Scientific and Technical Report on
"Advanced NMR Approaches in the Characterization of Coal"
DE-F22-90PC90290
Report No. 9
Gary E. Maciel
Department of Chemistry
Colorado State University
Fort Collins, CO 80523

DOE/PC/90290--9

DE93 006010

A considerable effort in this project during the past few months has been focussed on the development of ^1H and ^{13}C NMR imaging techniques to yield spatially-resolved chemical shift (structure) information on coal. In order to yield the chemical shift information, a solid-state NMR imaging technique must include magic-angle spinning, so rotating gradient capabilities are indicated. A ^{13}C MAS imaging probe and a ^1H MAS imaging probe and the circuitry necessary for rotating gradients have been designed and constructed. The ^1H system has already produced promising preliminary results, which are briefly described here.

One of the main problems with high-resolution (i.e., chemical shift) imaging with protons in solids is the necessity to average the ^1H - ^1H dipolar interactions by a multiple-pulse sequence (as demonstrated by Cory et al. with the CMG-48 sequence), without also averaging away the chemical shift (which, unfortunately, CMG-48 also does). Another important issue, which was a part of the impetus for the development of CMG-48, is to minimize the offset dependence of the line-narrowing capability generated by the magnetic field gradients applied in an imaging experiment. We felt that a TREV line-narrowing sequence would be advantageous for minimizing the offset dependence and accordingly developed a CRAMPS technique based on TREV.

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

gr

Line-narrowing effects of the ^1H TREV-8 pulse sequence applied to solids rotating at the magic angle were examined systematically. Polycrystalline samples of adipic, malonic and fumaric acids, adamantane, a barley seed and a pyridine-saturated coal sample were examined. The TREV-based ^1H CRAMPS technique was seen to provide reasonably effective line narrowing for both "soft" solids (adamantane) and rigid solids (carboxylic acids), enabling ^1H chemical-shift fine structure to be observed. For adamantane, the ^1H linewidth was reduced from about 13 kHz (static) to 40 Hz (CRAMPS) at a 840 Hz MAS rate and a pulse spacing (τ) of 7.8 μs . The line-narrowing efficiency increases for larger τ values, but does not approach the efficiency of BR-24. However, unlike the case for BR-24 or other multiple-pulse line-narrowing sequences based on the formation of solid echoes, the line-narrowing efficiency of TREV-8 is sustained for large offsets. This feature, along with a relatively large bandwidth, may be an important factor when implementing the sequence in ^{19}F CRAMPS and in solid-sample imaging studies. TREV-CRAMPS may also be useful for samples in which the somewhat inferior line-narrowing behavior may be masked by line-broadening effects not based on homonuclear dipolar interactions, e.g., in typical coal samples.

Having demonstrated that TREV is suitable for a CRAMPS experiment with solids like coal, we developed a proton NMR imaging experiment based on line-narrowing by TREV-CRAMPS with a time-dependent magnetic field gradient, and demonstrated it on a solid adamantane phantom. Spatial resolution similar to what is achieved in liquids NMR imaging, about 40 μm , was achieved with the magnetic field gradient applied as 16 μs pulses in the windows of the rf pulse sequence. The technique benefits from a favorable time average of the magnetic field gradient, relatively

wide windows between the magic-echo sandwiches, and the good off-resonance line-narrowing characteristics of the TREV technique. High-resolution chemical shift information is retained in this experiment, and hence potentially attainable. The low MAS rate, as compared to what is used in imaging experiments in which MAS is the only line-narrowing procedure used, may be of importance if one wishes to apply ^1H NMR imaging to certain types of samples whose integrity may be compromised by fast MAS.

This work has been summarized in two papers that have been submitted for publication. Copies are included as part of this report.

Manuscripts submitted

Marian L. Buszko, Charles E. Bronnimann and Gary E. Maciel, " ^1H CRAMPS Based on TREV," *J. Magn. Reson.*, submitted.

Marian L. Buszko and Gary E. Maciel, " ^1H Solid-State NMR Imaging by TREV-CRAMPS," *J. Magn. Reson.*, submitted.

*Preprints removed
and cycled separately.*

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

**DATE
FILMED**

2 / 5 / 93