PROGRESS REPORT

Ultrasensitive Fluorescence Detection of DNA

DE-FG03-91ER61125

Investigators: Richard A. Mathies and Alexander N. Glazer

Report Period: February 1, 1991 to January 31, 1992

Our research progress is summarized in the seven publications and three new patent filings listed below. Reprints of the recently published papers have been enclosed to describe the technical details. The basic accomplishments and their importance are described briefly below.

Publications 2 and 3 present a new method for performing capillary electrophoresis. Although capillary electrophoresis is a rapid separation method with excellent resolution, the throughput is low because only one capillary can be run at a time. To resolve this problem, we have been working on the production of capillary array devices. Dr. Huang, supported by a DOE Human Genome Postdoctoral Fellowship, has developed a new method for preparing capillary arrays and for detecting the fluorescent sequencing fragments. We have shown that very large numbers of capillaries can be detected with high sensitivity by scanning the planar array past a confocal detection system and sequentially detecting the signal from each capillary. Also, since capillary electrophoresis can be performed with high electric fields, the separation is complete in 1-2 hours. CAE is thus a high-speed, high-throughput DNA sequencing method. Our first publication (2) describes the separation of G-fragments on a four-capillary array. Single base resolution is observed out to 500 bases in less than two hours. This established the feasibility of using the confocal detection system to detect capillary arrays. Our second publication (3) presents the application of capillary arrays to DNA sequencing using a two-color detection system to detect four sets of sequencing fragments in only one capillary. To do this we developed a novel binary coding method that labels the four sets of sequencing fragments with unique ratios of primers labeled with two different fluorescent dyes. The development and advantages of this binary coding method are described in detail in this publication. This coding method and apparatus have been used to sequence M13 DNA on a 25 capillary array. The raw sequencing rate of this array is 6 kb/hr and it is projected that rates of up to 30 kb/hr should be feasible with such a system. We also published a Product Review in Nature that describes the use of capillary array electrophoresis and binary coding to perform high-speed, high-throughput DNA sequencing (6).

We have shown that a number of polycationic highly fluorescent dyes form complexes with double-stranded DNA (dsDNA) which are stable to electrophoresis and have characterized in detail such dsDNA complexes with TOTO (1,1'-(4,4,7,7-tetramethyl-4,7-diazaundecamethylene)-bis-4-[3-methyl-2,3-dihydro-(benzo-1,3-

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

thiazole)-2-methylidene]-quinolinium tetraiodide) and oxazole yellow dimer (YOYO; an analogue of TOTO with a benzo-1,3-oxazole in place of the benzo-1,3-thiazole). TOTO and YOYO are virtually non-fluorescent in solution, but form highly fluorescent complexes with dsDNA, up to a maximum dye to DNA bp ratio of 1:4, with >1000-fold fluorescence enhancement (1, 4). We have developed an assay using YOYO for the quantitation of single-stranded and dsDNA in solution applicable over a range of DNA concentrations from 0.5 to 100 ng per ml (5). The fluorescent dsDNA-dye complexes allow detection of dsDNA on agarose and acrylamide gels with picogram sensitivity. We have applied these complexes in multiplex mapping experiments for accurate sizing and quantitation of restriction fragments (4). In collaborative experiments with Alla Lishanskaya and Jasper Rine (Human Genome Center, LBL), we have shown that in gel shift experiments the stable dsDNA-dye complexes can be used to detect heteroduplex-MutS complexes with a sensitivity comparable to radioisotopic detection. Our review describing the properties and applications of the dsDNA-dye complexes appeared recently in Nature (7).

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

PLANS FOR THE NEXT BUDGET PERIOD

1. *High-throughput DNA sequencing using capillary arrays.* Yiwen Wang will continue the development of capillary array electrophoresis for high-speed, high-throughput DNA sequencing. In particular, we want to explore modified methods for loading the DNA sequencing sample that can be easily automated and that will require less DNA. We want to adjust the sequencing reaction chemistry and the gel chemistry to permit enhanced reading of more bases on a capillary. We will also explore the use of improved fluorescent dyes for labeling the primer that will give better separation in the two different color channels.

2. *High-Speed DNA mapping using capillary electrophoresis.* Steve Clark has set up a confocal fluorescence microscope and used it for the detection of high-molecular weight, double-stranded DNA fragments separated on non-denaturing gels in capillaries. The goal is to explore the capabilities of capillary arrays and confocal detection for the high-speed, high-throughput detection of mapping fragments. This system is now working and we are able to run 10 or more capillaries and produce complete separations in only 15-30 minutes. Our work in the next year will explore the optimization of the scanning detection system, the examination of a variety of fluorescent dyes for staining and detecting DNA separated on these columns, and the choice of the optimum separation medium. In addition, in the next period we will build a capillary array scanner dedicated to performing non-denaturing mapping separations.

3. *High-Density Capillary Arrays.* Adam Wooley is now beginning to explore the development of new methods for fabricating high density capillary arrays and for performing separations on capillaries that have much smaller internal diameters. We want to determine how many separations can be performed at one time and how fast these separations can be performed.

4. *Probe and method development for in situ hybridization.* We are finishing the construction of a stage-scanned confocal fluorescence microscope with digital photon counting detection. This apparatus should provide high signal-to-noise fluorescent images of chromosomes. First, the scanning and excitation parameters will be optimized using our fluorescence optimization theory. Then, the signal-to-noise of the confocal system will be critically compared with the capabilities of a cooled CCD detector on the same microscope. Finally, the microscope will be used to develop and assay the performance of new fluorescent probes for fluorescence in situ hybridization studies. The objectives are to develop probes sufficiently fluorescent to detect short regions of target DNA and to simplify the procedures used for in situ hybridization.

We have designed DNA hairpin probes which contain a region of self-complementary DNA and a region that is complementary to the target sequence. We have shown that such short probes (60-70

nucleotides) form stable complexes with TOTO, YOYO, and ethidium homodimer (EthD). We also plan to examine the use of biotinylated dsDNA-TOTO complexes as probes. We will study the stability of these complexes under various conditions required to obtain high stringency hybridization and will determine the parameters that govern the relative strengths of interactions of TOTO, YOYO, and EthD with double- and single-stranded DNA.

We anticipate that detection with our multiply labeled probes will be much more sensitive than that possible with the currently used fluoresceinated avidin and fluoresceinated anti-avidin antibody combinations:

5. *Applications of dsDNA-dye intercalation complexes.* The combination of the highly fluorescent dsDNA-intercalation dye complexes with the confocal scanning system permits detection of restriction fragments in the picogram range. This means that much smaller amounts of DNA are needed for such purposes as analysing restriction fragments of cosmids and ascertaining overlap between clones in contigs. To exploit this capability to analyze samples of unknown low DNA content (for example, such as those obtained by PCR amplification), we have developed a sensitive method for the quantitative determination of dsDNA concentration in solution. Reliable routine quantitation, utilizing conventional steady state fluorescence measurements, is possible down to 1 ng per ml. With this assay in place, we will explore the reliability of cosmid mapping and ordering by our procedure of running a standard mixture of dsDNA fragments stained with one dye and of unknown fragments stained with a second dye, both in the same lane.

In preliminary polyacrylamide gel shift experiments, we have shown that we can detect and quantitate complexes of dsDNA with transcription factors. Our procedures are both more sensitive and an order of magnitude more rapid than currently used procedures relying on radioisotopic labeling. We are completing the methodology development for the study of DNA-protein complexes and believe that it will find wide acceptance.

PRESENTATIONS SINCE THE LAST PROGRESS REPORT

1. "Capillary Array Electrophoresis", R. A. Mathies, invited presentation at the workshop on New Methodologies in Clinical Laboratories, annual meeting of the American Society for Human Genetics, San Francisco CA, November 11, 1992.
2. "Capillary Array Electrophoresis Using Confocal Fluorescence Detection", X. C. Huang, M. A. Quesada and R. A. Mathies, oral presentation at the Federation of Analytical Chemistry and Spectroscopy Societies (FACSS), Philadelphia PA, September 20-25, 1992.
3. "Capillary Array Electrophoresis Using Confocal Fluorescence Detection: An Approach to High-speed, High-throughput DNA Sequencing", X. C. Huang and R. A. Mathies, poster at the 1992 Gordon Conference on Analytical Chemistry, New Hampshire, August 2-6, 1992.
4. "Capillary Array Electrophoresis Using Laser-Excited Confocal Fluorescence Detection", X. C. Huang and R. A. Mathies, oral presentation at the symposium on high-performance electrophoresis at Science Innovation '92, San Francisco CA, July 21-25, 1992.
5. "Designer Fluorescent Probes Using Stable DNA-dye Intercalation Complexes", A. N. Glazer, oral presentation at the symposium on non-isotopic detection at Science Innovation '92, San Francisco CA, July 21-25, 1992.
6. "Fluorescent Methods in Novel Approaches to DNA Sequencing", R. A. Mathies and A. N. Glazer, LBL Human Genome Center Group Seminar, May 20, 1992.
7. "High-sensitivity Fluorescence Detection of DNA", R. A. Mathies, Joint Chemistry and Biochemistry Seminar, MIT, Cambridge MA, May 8, 1992.
8. "High-sensitivity Fluorescence Detection and the Human Genome Project", R. A. Mathies and A. N. Glazer, Structural Biology Seminar, Chemistry Department, U. C. Berkeley, April 27, 1992.

PUBLICATIONS SINCE LAST PROGRESS REPORT

1. Rye, H. S., Yue, S., Quesada, M. A., Haughland, R. P., Mathies, R. A. and Glazer, A. N., Picogram Detection of Stable Dye-DNA Intercalation Complexes with a Two-Color, Laser-Excited Fluorescence Gel Scanner, Methods in Enzymology, Recombinant DNA, Part H, Ed. R. Wu, 217, 414-431 (1993).
2. Huang, X. C., Quesada, M. A. and Mathies, R. A. Capillary Array Electrophoresis Using Laser-Excited Confocal Fluorescence Detection, Analytical Chemistry 64, 967-972 (1992).
3. Huang, X. C., Quesada, M. A. and Mathies, R. A. High-Speed, High-Throughput DNA Sequencing Using Capillary Array Electrophoresis, Analytical Chemistry 64, 2149-2154 (1992).
4. Rye, H. S., Yue, S., Wemmer, D. E., Quesada, M. A., Haughland, R. P., Mathies, R. A. and Glazer, A. N. Stable Fluorescent Complexes of Double-Stranded DNA with Bis-intercalating Asymmetric Cyanine Dyes: Properties and Applications, Nucleic Acids Research 20, 2803-2812 (1992).
5. Rye, H. S., Dabora, J. M., Quesada, M. A., Mathies, R. A. and Glazer, A. N. Fluorometric Assay Using Dimeric Dyes for Double- and Single-stranded DNA and RNA with Picogram Sensitivity, Analytical Biochemistry, in press.
6. Mathies, R. A. and Huang, X. C. Capillary Array Electrophoresis: An Approach to High-speed, High-throughput DNA Sequencing, Nature (London) 359, 167-169 (1992).
7. Glazer, A. N. and Rye, H. S. Stable Dye-DNA Intercalation Complexes as Reagents for High-Sensitivity Fluorescence Detection, Nature (London) 359, 859-861, 1992.

PATENT ACTIVITY SINCE LAST REVIEW

1. Laser-Excited Confocal Fluorescence Gel Scanner, K. Peck and R. A. Mathies, patent 5,091,652 issued February 25, 1992.
2. Multichromophore Fluorescent Probes using DNA Intercalation Complexes, A. N. Glazer, K. Peck and R. A. Mathies, patent filed March 14, 1990, Serial No. 07/493,307, pending.
3. High-Speed, High-Throughput DNA Sequencer using Capillary Array Electrophoresis and Fluorescence Detection, X. C. Huang, M. A. Quesada and R. A. Mathies, disclosed to U. C. Berkeley on 8/9/91 and filed on 2/24/92.
4. Multiple Tag Labeling Method for DNA Sequencing, X. C. Huang and R. A. Mathies, disclosed to U. C. Berkeley on 11/21/91 and patent filed ~3/30/92.
5. DNA Complexes with Dyes Designed for Energy Transfer as Fluorescent Reagents, A.N. Glazer and S.C. Benson, disclosed to U.C. Berkeley on 11/18/92.

END

DATE
FILMED
12/28/92