DOE/ER/54107-3
DEVELOPMENT OF A CHEMICAL KINETIC
MEASUREMENT APPARATUS AND
THE DETERMINATION OF THE REACTION
RATE CONSTANTS FOR
LITHIUM-LEAD/WATER INTERACTION

Status Report

October 1, 1991 - March 15, 1993

Dr. Paul Orleans Biney
Mechanical Engineering Department
Prairie View A&M University
Prairie View, Texas 77446

April 1993

PREPARED FOR THE U.S. DEPARTMENT OF ENERGY
UNDER GRANT NO DE-FG05-90ER/54107

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

MASTER

Se
DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

PREPARED FOR THE U.S. DEPARTMENT OF ENERGY UNDER GRANT
NUMBER DE-FG05-90ER54107

Technical Progress Report For the Period
October 1, 1991 - March 15, 1993

On the Development of a Chemical
Kinetic Measurement Apparatus and
the Determination of the Reaction Rate Constants
For
Lithium-Lead/Steam Interaction

Prepared by:

Dr. Paul Orleans Biney
Department of Mechanical Engineering
P.O. Box 397
Prairie View A&M University
Prairie View, Tx. 77446
Telephone: 409-857-4023
Fax: 409-857-2222

Submitted to the

OFFICE OF FUSION ENERGY
DEPARTMENT OF ENERGY

Through the

Prairie View A&M Research Foundation
P.O. Box 3578
College Station, Texas 77843
March 1993

Funded Project Period
09/21/90 - 3/21/93

PRINCIPAL INVESTIGATOR

Paul Orleans Biney, Ph.D.
Assoc. Professor
Mechanical Engineering Dept.

ABSTRACT

An experimental set-up for accurate measurement of hydrogen generation rate in Lithium-Lead ($\text{Li}_{17}\text{Pb}_{83}$) Steam or water interactions has been designed. The most important features of the design include a pneumatic actuated quick opening and closing high temperature all stainless steel valve used to control the reaction time and the placement of most measuring devices below a water line to minimize leakage of the hydrogen collected. A PC based data acquisition and control system provides remote process sequencing, acquisition and control of all major components of the set-up. Initial tests indicate that the first design objective of maintaining leakproof gas collection chamber has been achieved. Initial pressure tests indicated that the pressure drop over a time span of 30 minutes was within the tolerance of the pressure transducer used to measure the pressure (within 0.690 kPa) at a nominal system pressure of 685 kPa. The experimental system hardware, data acquisition and control programs and data analysis program have been completed, tested and are currently functional.

TABLE OF CONTENTS

	Page
ABSTRACT	i
LIST OF FIGURES	ii
LIST OF TABLES	iii
I. INTRODUCTION	1
A. Objectives of Research	1
B. Project Tasks	1
C. Graduate and Undergraduat Research Assistants	2
II. DESIGN OF EXPERIMENTAL SYSTEM	4
A. Formulation Experimental Test Matrix	4
B. Preliminary Determination of Hydrogen Gas Pressure in Gas Region of Upper Chamber	5
C. Conceptual System Design	6
D. Intermediate Experimental System Design.....	10
(i) Details of Lower Reaction Vessel	10
(ii) Liquid Metal Furnace	10
(iii) Upper Vessel Details.....	14
(iv) Mid-Section Butterfly Valve.....	14
(v) Design of Convection Coil.....	17
E. Initial Testing of Intermediate Experimental System	18
F. Final Experimental System Design	2
G. Operation of Final Experimental System	26
H. Lithium Lead Processing & Handling	27

III. DATA ACQUISITION AND CONTROL HARDWARE	29
A. Control of Solenoid Valves	34
B. Control of Liquid Metal Temperature	34
C. Control of Upper Water Temperature	35
D. Pressure Transducers	35
i. Upper Chamber Pressure	35
ii. Lower Gas Pressure	35
IV. CALIBRATION AND TESTING OF EXPERIMENTAL HARDWARE	37
A. Initial System Test	37
B. Calibration of Upper Gas Volume.....	37
C. Upper Chamber Pressure Test	40
V. DATA ACQUISITION PROGRAMS AND METHOD OF ANALYSIS	41
A. Data Acquisition and Control Program	41
B. Data Reduction and Analysis	46
VI. MODIFIED MASS TRANSFER MODEL FOR LITHIUM-LEAD WATER REACTION	53
A. Basic Assumptions	53
B. Model Equations	54
VII. PROBLEMS ENCOUNTERED AND STATUS OF RESEARCH	60
A. Problems Encountered	60
B. Status of Research	61
VIII. CONCLUSIONS	62

APPENDIX A	DETERMINATION OF THEORETICAL HYDROGEN PRESSURE	66
APPENDIX B	LISTING OF DATA ANALYSIS PROGRAM.....	68
APPENDIX C	LISTING OF DATA ACQUISITION AND CONTROL PROGRAM.....	81

List of Figures

Figure 1	Conceptual System Design.....	9
Figure 2	Details of Internal Cooling Channel in Lower Flange	11
Figure 3	Details of Lower Reaction Vessel.....	12
Figure 4	Details of Lower Furnace Design.....	13
Figure 5	Details of Upper Vessel.....	15
Figure 6	Pneumatic Controlled Butterfly Valve.....	16
Figure 7	Partial Assembly of Preliminary System.....	20
Figure 8	Final Design of Lower Vessel.....	22
Figure 9	Final Design of Upper Chamber.....	23
Figure 10a	Assembly of Final Experimental Set-Up.....	24
Figure 10b	Apparatus for Melting and Casting Molten Lithium-Lead	25
Figure 11	Photographs of System, Data Acquisition System and Control Panels ...	28
Figure 12	Data Acquisition and Control System.....	30
Figure 13	Simplified Schematic Control Diagram.....	33
Figure 14	Set-Up for Calibrating Upper Gas Volume.....	34
Figure 15	Simplified Flow Chart of Data Acquisition and Control Programs	47
Figure 16	Reference Coordinate System for Mass Transfer Reaction Model	55

LIST OF TABLES

	Page
Table 1 Final Experimental Test Matrix	7
Table 2 Theoretical Pressure of Hydrogen Generated in the $\text{Li}_{17}\text{Pb}_{83}$ /Water Reaction	8
Table 3 Design Specifications for Cooling Coil.....	19
Table 4 Analog Input/Output Features of Data Acquisition and Control System ...	31
Table 5 Digital Input/Output and Power Control Features of Data Acquisition and Control System	32
Table 6 Summary of Upper Chamber Calibration Results.....	39
Table 7 Summary of Controlled Equipment and Relays	42
Table 8 Summary of Thermocouples and Channel Connections	43
Table 9 Description of the Six Phases of Each $\text{Li}_{17}\text{Pb}_{83}$ /Water Reaction Experiment	44

I. INTRODUCTION

This status report covers work done on this project for the period October 1, 1991 through March 15, 1993. The greatest challenge in the design was ensuring a system with minimum leakage rate in order to obtain accurate accounting of reaction product gases. This greatly affects the placement of thermocouples, transducers and other instruments needed in the design.

A. Objectives of Research

The objectives of the research are:

- (i) The design and construction of an experimental set up that can be used to obtain accurate measurements of hydrogen generation rate during the initial and subsequent stages of liquid metal/water or steam reactions.
- (ii) The use of data collected in conjunction with an improved liquid metal transport reaction model developed to determine the empirical chemical kinetics reaction rate constants.
- (iii) To involve minority graduate and undergraduate students in this research.

B. Project Tasks

To enable the objectives listed above to be met, the overall project was divided into 10 main steps, namely

1. The search for qualified graduate and undergraduate students for the research and their subsequent introduction to the project.
2. Preliminary design of experimental system and determination of experimental test matrix.
3. Identification of manufactures/vendors and price list for all major components in preliminary design.
4. Modification of preliminary design. The final design of the experimental system.
5. Production of Working Drawings.

6. Procurement of components and modification of Liquid Metal Transport Reaction Model for experimental analysis.
7. Construction of experimental system.
8. Testing & correction of experimental system.
9. Data Collection on Hydrogen generation rate for liquid Lithium-lead/water reaction
10. Analysis of Data using the modified liquid metal Transport Reaction Model to obtain the empirical reaction rate constants.
11. Preparation of final Research Report.

The details of the work done from phase 1 through 8 of the project are described in this report.

C. Graduate and Undergraduate Research Assistants.

The first phase of the project involved obtaining qualified graduate and undergraduate students to work on the project. In this regard, James Akpan, an MS graduate student in the Mechanical Engineering Department was appointed a graduate Research Assistant for the project. In addition, Robert Tette, the most outstanding undergraduate Mechanical Engineering Junior student, expressed great interest in the research work and was hired as an Undergraduate Assistant in the project.

While James was primarily in charge of the overall research, his concentration was in the more technical aspects of data acquisition, analyses and chemical kinetics of the work. Robert was involved in the design of the system, procurement of components and assembly of the system. James and Robert spent a considerable portion of the first phase of the project in literature research On data acquisition, liquid metal/water/steam reactions, chemical reaction kinetics and fusion safety in general. In addition, Mr. Ali Ekhlassi was assigned to work on the

project as a Technician for the period June 1, 1992 to August 31, 1992 and Mr. Ali Ekhlassi was responsible for fabrication and assembly of the control panel, and testing of the experimental apparatus. Mr. Ekhlassi is an experienced Technician whose services was extremely beneficial to the project.

II. PRELIMINARY DESIGN OF EXPERIMENTAL SYSTEM

A. Experimental Test Matrix

To enable us determine the operating condition of the system to be designed, a detailed experimental matrix was initially developed. The matrix has been modified to enable only the most important parameters to be varied. The primary experimental variables are:

1. Initial Liquid Metal Temperature
2. Initial Water Temperature T_{uw}
3. Initial System Pressure P_{iug}
4. Total reaction time t_r and
5. Initial Mass of liquid Metal M_{lm} (de-emphasized)

To ensure that the surface area for reaction is constant throughout an experiment, the area was selected to be the maximum that prevents Taylor instability. The Secondary variables to be measured include:

1. Reaction product gas temperature T_{ug} .
2. Reaction product gas pressure P_{ug} .
3. Liquid Metal Temperature $T_{lm}(t)$
4. Lower gas chamber volume V_{lg} .
5. Low gas chamber pressure P_{lg} .

6. Lower gas chamber temperature T_m .

7. Initial Upper Gas Chamber Volume V_{iug} .

The final experimental test matrix for lithium lead/wate reaction is shown in Table 1

B. Preliminary Determination of Hydrogen Gas Pressure in the Gas Region of the Upper Chamber

In order to provide estimates of gas pressures and concentration of hydrogen gas in the gas region, a preliminary analysis was made to obtain the range of hydrogen gas concentration and partial pressures that may be encountered in the course of the experiment. The analysis was based on the following:

1. The mass of liquid metal initially in the lower reaction vessel is known.
2. The percent of metal reacted was varied from 5% to 100% (a typical range for such non-mixing reactions)
3. The upper chamber gas region volume was varied by varying the water level in the upper chamber (of diameter 3.81 cm) to produce gas volume heights of from 5 cm to 25 cm
4. Saturated liquid is assumed to be in the upper chamber, and the gas temperature was assumed to be 100°C.

The above information are used in conjunction with the chemical reaction equation for lithium-lead ($Li_{17}Pb_{83}$)/water reaction to determine the theoretical partial pressure of hydrogen generated during the reaction. The details of the calculation are given in Appendix A. Table 2 gives sample calculation results for ($Li_{17}Pb_{83}$)/water reaction. From the Table, it is evident that in addition to serving as a reactant and condensing medium for any water vapor generated as a result of the reaction, the water in the upper chamber also can be used to control the the partial

pressure of the hydrogen produced in the upper gas region by varying its column height. In doing so, we can also have control over the partial pressure of the hydrogen and therefore its concentration. These calculation results also provided us with the upper bounds of pressures to be expected in the system, and this information was used in designing for the minimum thickness of both the upper and lower vessels. Table 2 gives the summary of the results of the calculations, the details of which are given in Appendix A.

C. Conceptual System Design

Figure 1 shows the initial conceptual design of the setup. In this design, significant changes were made in the upper chamber and mid-section area of the apparatus. The attachment of the pressure transducers, thermocouples, and the convection coil have been located below the water level to minimize leakage of the hydrogen gas produced. The transducer and thermocouple wells have been modified by replacing them with a design that allows direct contact of the thermocouples with the gas and water regions. The convection coil is passed through a flanged part at the base of the upper chamber and below the water level. The principle behind these changes are that liquid leakages are far easier to control than hydrogen gas leakage. Thus we want to ensure that there are no flanged or screwed joints in the region where the gas collects after the reaction.

**Table 1 Final Experimental Test Matrix for
Lithium Lead/Water Reaction**

Liquid Metal Temperature °C	Initial Water Temperature °C	Reaction Time s	Liquid Metal Mass gm
350	90	5,10,20,200	60
350	70	5,10,20,200	60
400	60	5,10,20,200	60
400	90	5,10,20,200	60
400	90	5,10,20,200	30
500	90	5,10,20,200	60
500	70	5,10,20,200	60
500	90	5,10,20,200	30
600	90	5,10,20,200	60
600	70	5,10,20,200	60
650	90	5,10,20,200	60
650	70	5,10,20,200	60

Table 2 Theoretical Pressure of Hydrogen Generated for $\text{Li}_{17}\text{Pb}_{83}$ /Water Reaction

Gas Accumulation Height in Upper Chamber (cm)	Gas Pressure (kPa)			
	Percent Reaction			
	5 %	10 %	20 %	100 %
5	195.5	391.1	782.1	3910.6
10	97.8	195.5	391.1	1955.3
15	65.2	130.4	260.7	1303.5
20	48.9	97.8	195.5	977.7
25	39.1	78.2	156.4	782.1

Figure 1 Conceptual System Design

D. Intermediate Experimental System Design

Several changes were made in the conceptual design shown in Figure 1, the most important of which is the cooling method for the butterfly valve section. As a result of the high temperatures encountered by the lower reaction chamber and subsequently conducted to the butterfly valve, it is extremely important that the section close to the valve seat be effectively cooled. Provision was made for cooling the valve by incorporating an internal cooling channel in the lower flange connecting the butterfly valve to the lower chamber. The details of this design is shown in Figure 2. The detail design of the various components of the system is described below.

(i) Details of Lower Reaction Vessel

Figure 3 shows the details of the lower reaction vessel which was designed to contain the liquid lithium lead. The flange at the top was welded to the lower part of this section. The flange has an internal cooling channel just beneath the butterfly valve seat. Chilled water from a bath is circulated through this channel to maintain satisfactory temperatures at the valve section. This section is constructed of stainless steel. At the lower section, a 4 inch long NANMAC eroding tip fast response thermocouple is permanently inserted - The tip of this thermocouple is just 1.5 mm below the liquid metal pool and thus closely measures the liquid metal temperature. All basic dimensions of the lower reaction chamber are provided in Figure 3.

(ii) The Furnace

Figure 4 shows the details of the design of the furnace into which the lower section described in Figure 3 containing the lithium-lead is inserted, and heated to required liquid metal

All dimensions are in inches

Figure 2 Details of Internal Cooling
Channel in Lower Flange.

Figure 3 Details of Lower Reaction Vessel

All dimensions are in inches

Figure 4 Details of Lower Furnace Design

temperature. Parts A and B constitute a Watlow semi cylindrical high watt density ceramic fiber heater (750 Watts). Two such heaters are used in the furnace providing a total heating power of 1500 watts. Parts C and D are cylindrical ceramic fiber insulation modules used to insulate the lower section of the furnace. Part E is made of ceramic fiber blanket insulation. The outer shell of the furnace is made of stainless steel metal sheet.

(iii) Upper Vessel Details

The upper vessel, constructed with 316 stainless steel is shown in Figure 5. Initially, space V contains saturated water vapor, whose temperature is measured using an adjustable probe NANMAC right angle thermocouple (not shown) connected to port A, which also eventually measures the upper gas temperature once reaction is initiated. A second thermocouple connected to port C measures the temperature of the water in the upper chamber. A pressure transducer connected to port D, measures the system pressure in the upper chamber. A convection coil is connected to ports E and F, and hooked to a constant temperature bath and is used to control the initial temperature of the water. The upper chamber is bolted to the butterfly valve through a flange F.

(iv) Mid-Section Butterfly Valve

A pneumatic controlled 2-1/2 inch size 316 stainless steel butterfly valve, shown in Figure 6, separates the lower reaction chamber from the upper collection chamber. The valve is an ANSI class 300 high temperature valve. The valve was chosen due to its negligible leakage characteristic, sealing capacity and cycle time (250 ms). The upper and lower sections of the

Figure 5 Details of Upper Vessel

Figure 6 Pneumatic Controlled Butterfly Valve

apparatus are connected to the butterfly valve through a flange with grafoil gaskets used to prevent leakage. The flange connection also facilitates easy disconnection of the lower portion of the apparatus for the purpose of loading and unloading the liquid-metal into and out of the vessel in a glove box. To ensure that the butterfly valve and its actuator are operated within their specified temperature range, the valve section is water-cooled internally as described earlier.

(v) Design of Convection Coil

The experimental determination of the rate of hydrogen production during Lithium-Lead/Water reaction involves two distinct stages, the heating stage and the reaction stage. During the heating stage, the lower portion of the apparatus is electrically heated to melt the lithium-lead and to bring it to the required initial temperature T_{LM} . The reaction stage begins when the butterfly valve is opened and water from the upper chamber makes contact with the liquid metal in the lower section. This stage ends when the butterfly valve is closed at the end of a pre-determined reaction time, t_r . During the heating stage, it is important to ensure that both the butterfly valve and the electrical resistance heating elements are maintained within their stated operating temperature range. It is also important to ensure that right size heaters and adequate insulation is provided to obtain a reasonable heating time (30-45 minutes). These require that a heat transfer analysis be done at the early design stages of the system to ensure that the system under design will satisfy the required constraints. A two-dimensional finite element analysis has been used to analyze the heat transfer characteristics of the lower portion of the apparatus during the heating stage. During the heating stage of the experiment, heat is transferred from the lower chamber to the upper chamber, and subsequently by convection to the water in that chamber.

This heat has to be removed to maintain a constant water temperature in the upper chamber. The description below gives summary of analysis performed to ensure that the convection coil selected has adequate heat transfer area to remove this heat. Finite element analysis described earlier indicated a total of 375 Watts of heat will have to be removed. Based on this a number of constant temperature baths were reviewed to provide the cooling capacity and flow. Also the pump characteristics of the baths were noted. From the experimental test matrix given earlier in Table 1, the temperature of the pool of water outside the coil ranges from 60°C to 90°C. The above information, together with an assumed bulk inlet temperature of 5°C at the inlet of the coil were used to estimate the heat transfer area, the length of tube required, the head loss in the cooling coil circuit and the flow rate for the given system characteristics. The results of the coil design calculator are summarized in Table 3.

Figure 7 shows a partial assembly of the furnace, the lower reaction chamber, the mid-section pneumatic controlled butterfly valve, the cooling coil and the upper collection chamber.

E. Initial Testing of Intermediate Experimental System

In July 1992, the assembly of the experimental setup shown in Figure 7 was completed. The first test performed was the system pressure tests, and during this test, it was found that the system could not hold pressure. Systematic pressure test performed on the individual component

Table 3 Design Specifications for Cooling Coil

Tube Material	316 Stainless Steel
Tube Inside Diameter	3/16 Inch
Tube Outside Diameter	1/4 Inch
Minium Tube Length Required	28 Inches
Design Tube Length	40 Inches
Mean Coil Diameter	1 3/8 Inches
Bulk Fluid Inlet Temp	5°C
Flow Rate Through Tube	5 Liters/Min

Figure 7 Partial Assembly of Preliminary System

indicated a welding imperfection on the cooling coil. It was also found that the butterfly valve had a higher leakage rate than specified. The failure of the butterfly valve to perform to specification was a big blow to the progress of the research. A representative of the valve manufacturer was brought on sight and upon inspection, found the valve to be defective. The manufacturer recommended the use of a zero-leak all stainless steel Vitron ball valve. It was also necessary to change the valve size from 2 1/2" to 1" due to the high cost of the 2 1/2" Vitron ball valve. This valve was ordered to be manufactured in August 1992. The new valve was not received until December, 1992. Upon inspection, it was found that a wrong valve stem length was attached. It was returned to the manufacturer. The final valve with the right specification was received in January 15, 1993. As a result of a change in valve size from 2 1/2 inches to 1 inch, several modifications had to be made in both the lower and upper vessels.

F. Final Experimental System Design

A new upper vessel was designed and fabricated to match the new valve between January 15th and March 1, 1993. Several modifications were made in the lower vessel to fit the new valve. Figure 8 shows the details of the modified lower vessel. Due to the reduction in size of the valve and its flanges the details of the internal cooling was modified as shown in the Figure 8. Figure 9 shows the details of the final upper chamber design. The enlarged section previously on top of this chamber was eliminated to reduce fabrication cost and for simplicity.

All the good leakage prevention features in the preliminary design were retained in this final design. The new experimental design has been assembled and all instrumentation attached. Figure 10a shows a schematic assembly of the final experimental system. Figure 10b shows

Figure 8 Final Design Details of Lower Vessel

Figure 9 Final Design Details of Upper Vessel

Figure 10a Assembly of Final Experimental System

Figure 10-b Apparatus for Melting and Casting Molten Lithium-Lead

photographs of the system, the control panel and the data acquisition and control systems when they were being assembled.

G. Operation of Final Experimental System

Figure 10a shows a partial assembly of the system. The furnace sections, the lower section, and the upper sections have been previously described. During the liquid metal heating phase of the experiment, the furnace is on, and the sample of lithium lead in lower reaction chamber R is heated to a predetermined temperature. The main ball valve is closed and argon flows through the Argon Inlet solenoid valve (AIV) through the lower chamber and exits through the argon outlet solenoid valve (AOV). The argon purge valve (APV), is also opened and the system pressure in the lower chamber monitored by the lower gas pressure transducer (LGP) and temperature by lower gas temperature thermocouple (LGT). A 60 seconds count down time is planned prior to the opening of the main ball valve. At the beginning of this count down period, valves AIV and AOV are closed to stop the argon flow at which time pressure transducer LGP and thermocouple LGT monitor the pressure and temperature of the closed lower argon space in the lower chamber. Data obtained during this countdown phase is used to determine the mass of argon in the lower chamber. At the end of the countdown period, valve APV is closed, the main ball valve (MBV) is opened and reaction between the lithium-lead and water is initiated. The reaction is made to continue for a pre-determined period of time (t_R) after which the butterfly valve is closed, Valve AOV opened and APV closed. Any hydrogen produced after that time is made to flow into a collection tank (HCT). During the reaction period, the hydrogen produced collects above the water in the upper chamber, where the temperature and

pressure are measured by UGT and UGP respectively. The timing, sequencing and control of all the solenoid valves and the pneumatic controlled butterfly valve are done through Keithly Measurement and Control System Model 575-2.

H. Lithium Lead Processing and Handling

Lithium-lead metal was received in December 1992 from Oakridge National Lab in a 1 gallon container. In March 1993, a heater and valve system were assembled on the lithium lead container. Figure 11 shows the schematic design of the melting system. We attempted to melt and cast the metal into small discs under argon atmosphere, ready to be used in the experiments. The metal was heated to about 400°C and after casting some few discs, the valve failed and the liquid metal poured into the glove box uncontrollably. This damaged the rubber gloves enabling air to enter the chamber, causing the molten metal to be contaminated.

We are currently awaiting for new supply of the lithium lead from Oak Ridge National Lab to continue and complete the work.

Figure 11 Photographs of System, Data Acquisition
and Control Panels

III. DATA ACQUISITION AND CONTROL HARDWARE

In this chapter, the hardware used for the acquisition of data and control of the experiment are described.

At the heart of the data acquisition and control system is a Keithley Model 575-2 Measurement and Control unit shown in Figure 12. A summary of the analog input/output and digital input/output features of the unit are given in Tables 4 and 5 respectively. Sixteen of the 32 digital input/output channels are dedicated for power control of on/off devices including all solenoid valves and the ball valve through the use of a power control module and relay board. The power control module (PCM-3) is a general purpose remote relay card for the control of power AC and DC devices. It permits direct interface of relays, heaters, meters, actuators, and other AC and DC devices over the full range of voltages from 10 to 280 V. A 16-channel thermocouple module capable of handling J,K,S,T,B,E and R single or mixed thermocouples and having an iso-thermal block is used for all temperature measurements. Figure 13 Shows the simplified schematic control diagram. The solid state relays SSR-0 through SSR-6 are operated through software, (the Data Acquisition and control program (DACP) developed for this work) and serve as the main on-off switches for all equipment.

Figure 12 Data Acquisition and Control System

Table 4 Analog Input/Output Features of Data Acquisition and Control System

Feature	Description/Range
SPEED	50,000 READINGS/S
RESOLUTION	16 BITS
FULL SCALE RANGES	+100mV, +200mV, +500mV +1V, +2V, +5V, +10V
CHANNELS	8 DIFFERENTIAL 16 SINGLE ENDED PLUS 8 ADDITIONAL SINGLE ENDED
ANALOG OUTPUT	
CHANNELS	2, SINGLE ENDED
OUTPUT RANGES	+10V, +5V, +2V, +1V
RESOLUTION	13 BITS

Table 5 Digital Input/Output and Power Control Features of Data Acquisition and Control System

Feature	Description/Range
CHANNELS	32 NON-ISOLATED, PROGRAMMABLE FOR INPUT OR OUTPUT IN GROUPS OF 8 CHANNELS. 16 CHANNELS CAN DRIVE POWER CONTROL RACK FOR ON/OFF CONTROL OF POWER DEVICES.
TRIGGER FUNCTIONS	
CHANNELS	1, DIFFERENTIAL INPUT
TRIGGER SOURCE	EXTERNAL INPUT, ANY INPUT CHANNEL, OR SOFTWARE STORBE.
RANGES	0 TO -10V, 0 TO -IV, 0 TO 10V 0 TO -10V
RESOLUTION	8 BIT
INPUT COUPLING	AC OR DC

Figure 13 Simplified Schematic Control Diagram

A. Control of Solenoid Valves

Relays SSR-2 through SSR-6 control the on-off states of the cooling coil/flange circulating water (CWV), the argon inlet valve (AIV), the argon outlet valve (OAV), the upper-purge valve (UPV), and the main ball valve (MBV) respectively. The sequencing timing and duration of operation of these valves are done through DACP.

B. Control of Liquid Metal Temperature

Liquid metal thermocouple (LMT) reads the liquid metal temperature. LMT terminates on a terminal strip from where two sets of thermocouple extension wires take off, one set to a temperature controller (Temp. Cont.-1) and the other set to the temperature measurement module in the 575-2 DAC System. The preset liquid metal temperature is programmed into Temp. Cont.-1, and this controller uses the temperature output of the LMT to control the liquid metal heater LMH, and thus the liquid metal temperature. LMH could not be directly connected to SSR-1 due to the high current pulled by the heater. High line current solid state relay (SSR-10) was initially connected between SSR-1 and the heater, thereby enabling SSR-1 to energize SSR-10 to turn the LMH on. That arrangement did not work due to the leakage current of the two solid state relays SSR-1 and SSR-10. It became necessary to install a mechanical relay MR-2 between SSR-1 and SSR-10. The signal to energize SSR-1 is obtained from the temperature reading on the lower furnace thermocouple (LFT) which reads the heater surface temperature that has been preset to protect the heater element from overheating. As long as the experiment has been initiated and LFT is below the setpoint, the DAP energizes SSR-1 which in turn energizes MR-2 and MR-2 energizes SSR-10. The dynamic control of liquid metal temp through

on/off power regulation of LMT is done by the programmable temperature controller Temp. Cont.-1.

C. Control of upper water temperature

The control of the upper water temperature is identical to the control of the liquid metal temperature. The solid state relays involved are SSR-0, MR-1, SSR-9 and temperature controller T-C2. The thermocouples involved are the upper water thermocouple (UWT) and the upper furnace thermocouple (UFT)

D. Pressure Transducers

Two pressure transducers are used to measure the major pressures of the system. As shown in Figure 10-a.

(i) Upper Chamber Pressure

The upper chamber system pressure is the pressure of the gas mixture above the water level in the upper chamber and it is measured by a high accuracy Setra absolute pressure transducer (UGP). The 0-5 VDC linear output of the transducer is connected to the 575-2 DAC system and converted to kilopascals by DACP. This transducer requires 15-24 Volts DC Supply for excitation.

(ii) Lower Gas Pressure

The lower gas pressure is the pressure of the argon gas in the space between the main ball valve

and the liquid metal as shown in Figure 10-a. This pressure is used in conjunction with the temperature of the gas in that chamber and the total volume of the chamber to calculate the amount of argon gas initially in the system. This pressure is measured by a Setra absolute pressure transducer (LGP). The 0-5 VDC linear output is also converted to kilopascals by the DACP through the 575-2 DAC system.

IV. CALIBRATION AND TESTING OF EXPERIMENTAL HARDWARE

A. Initial System Test

The initial system tests performed include check of continuity of all electrical connections, testing of all instruments with and without the use of data acquisition and control system, calibration of upper and lower vessel volumes.

B. Calibration of Upper Gas Volume

In order to obtain an accurate measurement of the gas space above the water in the upper chamber, it was necessary to accurately calibrate or calculate the volume and to determine the mean and standard deviation of the measurement for statistical analysis of the final experimental data. Figure 14 shows the experimental set up for calibration of the upper gas volume. With the upper solenoid valve (USV) closed, upper pressure transducer (UGP) and upper gas thermocouple (UGT) removed, the upper chamber is filled (through valves A and B with water until the system is completely filled. The chamber is slowly drained through valves A and B into a finely graduated cylinder. The volumes and mass of water collected at different water levels on finely graduated indicator on the upper chamber are recorded. The measured results are then compared to the calculated results obtain from dimensions of the vessel and tubes. Table 6 summarizes the results of the upper chamber calibration tests. The experimental values of the volumes agree well with those calculated using vessel and tube dimensions. In the useful water level range of 0-9 inches, the maximum standard deviation is

Figure 14 Setup for Calibrating Upper Gas Volume

Table 6 Summary of Upper Chamber Calibration Results

Water Level Mark Inches	Expt 1 Volume CC	Expt 2 Volume CC	Expt 3 Volume CC	Calculated Volume CC	Mean	Standard Deviation
1	29.0	28.5	28.5	27.93	28.48	0.14
2	57.5	58.0		56.88	57.26	0.25
3	86.0	86.5	86.5	85.85	86.21	0.09
4	114.5	115.5	115.1	114.80	114.98	0.14
5	144.0	144.0	143.1	143.76	143.84	0.03
6	172.5	173.0	172.5	172.72	172.68	0.04
7	201.5	202.0	201.5	201.68	201.67	0.04
8	230.0	230.5	230.0	230.64	230.29	0.08
9	259.5	260.5	258.5	259.60	259.53	0.50
10	285.0	286.0	286.3	288.55	286.46	1.68
11	315.5	316.5	315.5	317.51	316.25	0.69
12	346.0	346.0	344.0	346.47	345.62	0.91
13	375.5	375.5	373.0	375.43	374.86	1.15

0.5 cc. Thus all upper gas volumes will be repeated as $xxx.xx + 0.50$ cc. The total volume of the space in upper chamber was measured to be 1386 cc. The calculated value obtained from system dimensions is 1388 cc. The volume used in all calculations is the average of the two, $1387.6 \text{ cc} + 1 \text{ cc}$.

The total volume of space in the lower chamber below the main ball valve (when in closed position) was to be 114.2 cc and the calculated value is 114.3 cc. This volume is repeated as 114.25. The total volume of space in the lower chamber with the ball valve opened (volume in ball included) is 137.1 cc.

C. Upper Chamber Pressure Test

Two pressure tests were performed on the upper chamber to check for leakages and the ability of the vessel to hold pressure. In the first test, with the ball valve closed the upper vessel was pressurized to 650 kPa. The pressure was monitored for three hours. The total pressure drop during this period was 5.5 kPa.

In the second test, the upper chamber was filled with water to a level just below the purge line and the system pressurized to 650kPa. After three hours, the pressure drop was 0.52 kPa, well within the tolerance of the pressure transducer accuracy. From the pressure test, we were satisfied with the ability of the experimental set-up hold pressure over extended period of time.

V. DATA ACQUISITION PROGRAMS AND METHOD OF EXPERIMENTAL ANALYSIS

A. Data Acquisition and Control Program

A brief description of the important features of a basic program for acquisition and control of the experiment is described in this section. The important equipment to be controlled are connected to relays on a Keithley PCM3 Relay Board. Table 7 gives the description of the relay connections. Seven thermocouples (TC) are installed in the system and directly connected to the thermocouple module (AIM7). Two extension wires carry signals from the liquid metal TC temperature and the upper water temperature TC to two temperature controllers. Table 8 gives a summary of thermocouples and channel connections. The two setra pressure transducers are directly connected to the AMM2 master module. Each experiment is divided into six phases, and Table 9 gives the summary of the important features of these phases.

Phase 1

During this phase, the upper water, the liquid metal heater and the circulating bath are turned on to heat the water in the upper chamber to a pre-determined temperature at which time a vacuum pump is used to control the pressure to obtain a saturated liquid in the upper chamber.

Phase 2

This phase involves activation of the liquid metal heater and the liquid metal heated to a predetermined temperature.

Table 7 Summary of Controlled Equipment and Relays

Relay Number	Input/Output Name	Description of Equipment Attached
0	UWH	Upper water heater
1	LMH	Liquid metal heater
2	CWV	Circulating water valve
3	AIV	Argon inlet valve
4	AOV	Argon outlet valve
5	USV	Upper solenoid valve
6	MBV	Main ball valve

Table 8 Summary of Thermocouples and Channel Connections

Description	Input/Output Names	Channel Number	Thermocouple type
Upper Gas Temp.	UGT	1	E
Upper Water Temp.	UWT	2	K
Upper Water Temp-C	Controller-2	-	K
Lower Water Temp.	LWT	3	E
Lower Gas Temp.	LGT	4	K
Liquid Metal Temp.	LMT	5	K
Liquid Metal Temp-C	Controller-1	-	K
Upper Furnace Temp.	UFT	6	K
Lower Furnace Temp.	LFT	7	K

Table 9 Description of the Six Phases of
Each Lithium-Lead/Water Reaction Experiment

Phase #	Phase Description	Time Interval Between Readings (Seconds)	No. of Data Acquired (Maximum)	Data	
				Saved for Analysis	Not Saved
1	Upper Water Heating	20	100		x
2	Liquid Metal Heating	20	150		x
3	Water & Metal Temp. Control	20	90		x
4	Count down to Reaction	1	60	x	
5	Chemical Reaction	.2	100	x	
6	Equilibrium Period	20	100	x	

Phase 3

This phase involves controlling the upper water and the liquid metal temperatures until they are within a given tolerance of their preset values. This step was found necessary for repeatability of the experiments to ensure an accurate generation of the reaction rate curves from different experiments with identical values of the liquid metal & upper water temperatures. When these two temperatures are within their proper range, phase 4 is activated automatically by the DACP.

Phase 4

This is the count down phase. There is a 60 seconds countdown, during which all pressures and temperatures are read every 1 second and stored later. The average values of each set of reading is used to establish the initial conditions in the system. Particularly, the temperature and the pressure of the argon gas in the lower chamber, which is completely closed at this time, is used to determine the amount of argon in the system. At the end of the 60 seconds countdown, the fifth phase is initiated.

Phase 5

This phase, the reaction phase, begins with the automatic opening of the main ball valve, to initiate the reaction. The valve remains open for a preset reaction time period after which the valve closes, indicating end of reaction. All temperature and pressure readings are recorded and saved during the reaction phase.

Phase 6

Phase 6 involves recording of pressure and all temperatures in the upper chamber for a period of 10 to 20 minutes or until the temperature in the chamber is uniform. The data at this phase is used to obtain the average value of the total amount of hydrogen generated during the reaction time period. Figure 15 gives a simplified flow chart of the Data Acquisition and Control Program. The complete program is given in Appendix C.

B. Data Reduction and Analysis

The data collected from the experiment, the physical inside dimensions of system, the leakage characteristics of the system

and information on the solubility of Argon and hydrogen in water are used to determine the amount of hydrogen generated

during the reaction. This is done by accounting for all calculable components of pressure in the system.

(i) Mass of Argon before Reaction

The mass of argon in the system just before the reaction can be calculated from the equation

$$N_{Ar} = \frac{(P_{wg} - P_{H_2O})}{RT_{wg}} V_{wg} + \frac{P_g V_{gc}}{RT_g} + N_{Ar,ini} (T_{wg} P_{wg}) \quad (1)$$

Figure 15 Simplified Flow Chart of Data Acquisition and Control Program

Figure 15 Simplified Flow Chart of Data Acquisition and Control Program

$$N_{H_2}=0 \quad (2)$$

The amount of Argon in solution is calculated using information on the solubility of argon in water at 25°C and 1 atm, and the fact that the solubility reduces to zero at $T_{sat} = 100^\circ\text{C}$ and 1 atm. Thus by interpolation, the solubility of argon, X_{Ar} is given by

$$X_{Ar}(T_w, P_{iug}) = [X_{Ar}(25^\circ\text{C}, 1\text{atm}) * \frac{(T_{sat}(P_{iug}) - T_w)}{(T_{sat}(P_{iug}) - 25)} \frac{P_{iug}}{1\text{atm}}] \quad (3)$$

where

T_w is the water temperature and

N_{H2O} is the amount of water in the upper chamber.

P_{iug} is the total initial system pressure in the upper gas region,

P_{H2O} is the vapor pressure of water at upper gas temperature

T_{iug} is the initial upper gas temperature

R is the universal gas constant,

- P_{lg} is the pressure in the lower chamber just before the reaction,
- T_{lg} is the temperature of the argon in the lower chamber just before reaction,
- V_{lgc} is the volume of argon in the lower chamber,
- V_{iugc} is the initial volume of the upper gas chamber and
- $N_{Ar(sol)}(T_w, P_{iug})$ is the amount of argon in the solution in the upper chamber at the given system pressure and upper water temperature.
- $T_{sat}(P_{iug})$ is the saturation pressure evaluated at the initial upper gas pressure.

The amount of argon in solution is then calculated using

$$N_{Ar}(T_w, P_{iug}) = [X_{Ar}(25^\circ C, 1atm) * \frac{(T_{sat}(P_{iug}) - T_w)}{(T_{sat}(P_{iug}) - 25)} \frac{P_{iug}}{1atm}] * N_{H_2O} \quad (4)$$

where N_{H_2O} is the amount of water in the upper chamber.

(ii) Mass of Argon and Hydrogen after initiation of Reaction

Once the reaction has been initiated, the mass of argon at any time is given by

$$N_{Ar} = (N_{i_{Ar}} - N_{Ar_{red}}) \quad (5)$$

The argon pressure is then given by

$$P_{Ar} = \frac{(N_{i_{Ar}} - N_{Ar_{red}})RT_{ug}}{V_{ugc}} \quad (6)$$

From the above information, the partial pressure of hydrogen is given by

$$P_{H_2} = P_{ug} - P_{H_2O} - P_{Ar} \quad (7)$$

The mass of hydrogen generated is then calculated using

$$N_{H_2} = \frac{P_{H_2} V_{ugc}}{RT_{ug}} + N_{H_{2sol}} \quad (8)$$

Where $N_{H_{2sol}}$ is the amount of hydrogen generated that goes into solution and is determined in a similar manner as the determination of amount of Argon in Solution. $N_{H_{2sol}}$ is calculated using

$$N_{H_{2sol}} = [X_{H_2}(25^\circ C, 1atm) * \frac{(T_{sat}(P_{ug}) - T_w)}{(T_{sat}(P_{ug}) - 25^\circ C)} \frac{P_{ug}}{1atm}] N_{H_2O} \quad (9)$$

(iii) Gas Leakage Corrections

Even though system pressure tests indicate negligible leakage, it may be necessary to correct the amounts of hydrogen accounted for in the upper chamber and in solution for leakage in order to accurately ascertain the amount of hydrogen produced from the reaction. Before each experiment, a pressure test is performed to determine the leakage rate of the system. A curve-fit is obtained for the pressure test data and used to determine the leakage rate using the following

equations. The total amount of argon leakage from time 0 to t_f is given by

$$N_{Ar_{leak}}(t) = \sum_{i=1}^n (t_i - t_{i-1}) * \left[-\frac{dP(t_i)}{dt} \right] * \frac{P(t_i)_{Ar}}{P(t_i)_{ug}} \quad (10)$$

Similarly, the amount of hydrogen leakage from time 0 to t_f is also given by

$$N_{H_2_{leak}}(t) = \sum_{i=1}^n (t_i - t_{i-1}) * \left[-\frac{dP(t_i)}{dt} \right] * \frac{P(t_i)_{H_2}}{P(t_i)_{ug}} \quad (11)$$

The corrected total amount of hydrogen generated from time $t=0$ to t_f is then given by

$$N_{H_2_{Total}}(t) = \frac{P_{H_2}(t)V_{ug}}{RT_{ug}} + N_{H_2_{gen}}(t) + N_{H_2_{leak}}(t) \quad (12)$$

and the argon inventory at any time given by

$$N_{Ar_{Total}}(t) = \frac{P_{Ar}(t)V_{ug}}{RT_{ug}} + N_{Ar_{gen}}(t) + N_{Ar_{leak}}(t) \quad (13)$$

The equation for the inventory of Argon is used to determine the accuracy of pressure

accounting and repeatability of the experiment. A leakage test on the system indicate the leakage correction can be neglected without loss of accuracy. A complete Fortran program that reads the raw experimental data and calculates the hydrogen generation as a function of time is given in Appendix B.

VI. MODIFIED MASS TRANSFER MODEL FOR LITHIUM-LEAD/WATER REACTION

A liquid metal transport reaction model developed by Herzog [1] has been improved and being used to estimate the empirical chemical reaction rate and diffusion parameters for the lithium-lead/water chemical reaction. This model is based on the premise that the rate of reaction during lithium-lead/water interactions is primarily controlled by the rate of diffusion of lithium atoms to and products away from the liquid metal surface right from the beginning of the reaction. One of the most difficult aspects of using the model is the requirement of determining the initial hydrogen generation rate. The experimental data during the first few seconds from the present research will provide reliable data for determining the initial hydrogen generation rate for use in the liquid metal transport model. This model proposed by Herzog [1], is discussed in detail by Biney et al [2].

A. Basic Assumptions

The liquid metal transport reaction model is based on the following assumptions: (1) the reaction occurs only at the surface, (2) the system is assumed to be one dimensional since the driving force of diffusion will result in an axial variation of molar concentration of reactants and products, (3) the concentration of lead is assumed to be approximately constant throughout the interaction since there are about 5 atoms of lead to 1 atom of lithium in the alloy and lead is chemically inert during the interaction, (4) the liquid metal is incompressible, (5) the gases

produced during the reaction are ideal, (6) there is no convective motion in either the gas or liquid metal pool and (7) there is no bulk mixing within the liquid metal pool. In the tests, the diameter of the liquid metal surface was stable relative to the Taylor unstable wavelength and thus the reaction surface area remains approximately constant. Figure 16 shows the reference coordinate system used for the model.

In the figure, the interaction surface is shown located a distances from the bottom of the liquid metal pool. Hypothetical concentrations of the products and reactants are also indicated on the Figure. A vapor film of thickness δ is shown above the interaction surface and a pool of subcooled water above the vapor film.

B. Model Equations

The basic equations are reviewed here. A complete description can be found in References 2 and 3.

Continuity Equations

For a multi-component system the continuity equation is given [5] by:

$$\frac{\partial C_i}{\partial t} = -\frac{\partial}{\partial z}(C_i V + J_i) \quad (15)$$

which when applied to the reactant and product results in the following two equations:

$$\frac{\partial C_u}{\partial t} = D_{im} \frac{\partial^2 C_u}{\partial z^2} - C_u \frac{\partial V}{\partial z} - V \frac{\partial C_u}{\partial z} \quad (15)$$

Figure 16 Reference Coordinate System for Mass Transfer Reaction Model

$$\frac{\partial C_{pd}}{\partial t} = D_{lm} \frac{\partial^2 C_{pd}}{\partial z^2} - C_{pd} \frac{\partial V}{\partial z} - V \frac{\partial C_{pd}}{\partial z} \quad (16)$$

The boundary conditions for the continuity equations are :

$$\text{At } z=0, \frac{\partial C_{Li}}{\partial z}=0; \text{ at } z=s, C_{Li}=0 \quad (17)$$

and the initial condition is :

$$C_{Li}(z,0)=C_{Li_{max}}. \quad (18)$$

Under assumptions 3 and 4, the equation of energy for the liquid metal pool is given by the equation

$$\rho \frac{DH}{Dt} = -\frac{\partial}{\partial z} (k_{lm} \frac{\partial T}{\partial z} + M_r H_r J_r) + \frac{DP}{Dt} \quad (19)$$

Substituting the relation $H = U + PV$ and simplifying, we obtain the equation

$$\rho c_p \left(\frac{\partial T}{\partial t} + V \frac{\partial T}{\partial z} \right) = k_{lm} \frac{\partial^2 T}{\partial z^2} - \left((J_L M_L c_{pL} + J_{Pd} M_{Pd} c_{pPd}) \frac{\partial T}{\partial z} + M_L H_L \frac{\partial J_L}{\partial z} + M_{Pd} H_{Pd} \frac{\partial J_{Pd}}{\partial z} \right) \quad (20)$$

where

$$J_i = -D_{lm} \frac{\partial C_i}{\partial z} \quad \text{and} \quad D_{lm} = D_o \exp\left(-\frac{\Delta E_d}{RT}\right) \quad (21)$$

and where C , V , J , M , D_{lm} , T and P are the molar concentration, velocity, molar flux, molecular weight, liquid metal diffusion coefficient, temperature and pressure respectively.

The boundary conditions for the energy equation are: $T(0, t) = T_c(t)$, a known temperature at the base of the metal pool; at $z=s$, $q_{cond} + q_{gen} = q_{conv}$ where q_{cond} is the heat transfer by conduction from the base to the surface, q_{gen} is the heat generated at the surface as a result of the exothermic reaction and q_{conv} is the heat transfer by convection from the reaction surface to the vapor film given respectively by:

$$\begin{aligned} q_{cond} &= -k_{lm} \frac{\partial T}{\partial z} \\ q_{gen} &= -R_L (n_{Pd} h_{Pd} - n_{react} h_{react}) \\ q_{conv} &= h_p (T_s, t) - T_{sat} \end{aligned} \quad (22)$$

The stable film boiling heat transfer coefficient h_{fb} is obtained from theory of film boiling on flat horizontal surface. Collier [3] gives the following equations for calculating this coefficient.

$$h_{fb} = h_c + 0.75h_{rad} \quad (23)$$

where

$$h_c = 0.62 \left[\frac{g(\rho_f - \rho_g)\rho_g k'_f i'_{fg}}{\lambda_c \mu_g (T_g - T_{sat})} \right]^{\frac{1}{2}} \quad (24)$$

and

$$\lambda_c = 2\pi \left[\frac{\sigma}{g(\rho_f - \rho_g)} \right]^{\frac{1}{2}} \quad (25)$$

is the unstable Taylor wavelength. The corrected latent heat of vaporization i'_{fg} , is given by:

$$i'_{fg} = i_{fg} \left[1 + 0.68 \left(\frac{c_{pg} \delta T}{i_{fg}} \right) \right] \quad (26)$$

The two continuity equations and the one energy equation constitute three equations with five unknowns: $T(z,t)$, $C_{Li}(z,t)$, $C_{pd}(z,t)$, D_o and E_d . Assumptions 2 and 3 imply that the molar flow rates of lithium and product ($LiOH$) are equal but opposite in direction. Thus the sum of the molar concentrations will be a constant equal to the initial lithium concentration in the alloy $C_{Li,max}$. Thus equation (16) can be replaced with

$$C_{Pd}(z,t) + C_{Li}(z,t) = C_{Li_{max}} \quad (27)$$

Finally the model is made to match two other conditions obtained from experimental results, namely the initial rate of hydrogen production, and a steady state amount of hydrogen generated t_f seconds into the reaction. These last two conditions bring the set of equations to 5 with 5 unknowns. The initial rate of hydrogen generation S , is calculated using experimental data

$$\frac{dN_{H_2}}{dt} = S \quad (28)$$

The experimental results will be used to determine the moles of hydrogen produced at the end of t seconds. The initial rate of hydrogen generation will be determined from the experimental data. These two pieces of experimentally derived information will be supplied to the liquid metal transport reaction model and the reaction rate constants B and E in the parabolic rate equation

$$\frac{\partial}{\partial t} \left[\frac{m}{A} \right] = B \exp \left[-\frac{\Delta E}{RT} \right] \quad (29)$$

as well as the constants in the Arrhenius equation D_0 and ΔE_d determined. One should note that mechanistically these constants B and ΔE are directly related to D_0 and ΔE_d in the liquid metal diffusivity, Equation (6).

VII. PROBLEMS ENCOUNTERED AND STATUS OF RESEARCH

A. Problems Encountered

Two major problems were encountered during the first six months of the project, all being the result of the project contract date occurring almost a month after beginning of the semester. The principal investigator was unable to obtain release time to actively work on the project in the Fall Semester of 1990 (August 28 - December 31, 1990) because of the late contract date. A college technician could not be assigned to work on the project in the Fall Semester of 1990 as a result of the late contract date. We could also not get the college technician to work on the project in the Spring Semester of 1991 (January - May 1991) due to heavy work load already being assigned by the college. In the summer of 1992, a technician was assigned to the project. By this time all the design and equipment acquisition had been completed. The complete system (including lower and upper chambers, all, heaters solenoid valves, butterfly valve and the data acquisition system) was assembled. In July 1992, pressure leakage tests was conducted on the vessels and the butterfly valve was found to be leaking in the closed position. Also the cooling coil had defective welding joint. Upon discussion with the valve manufacturer, the butterfly valve was returned, and a zero leakage all metal valve was ordered in August 1992 for replacement. The valve size for this valve had to be changed to 1" due to cost. The valve was not received until January 15, 1993. The change in the valve size necessitated redesign of the upper chamber and modifications in the lower chamber. The modifications and assembly of the

new system were completed in February 15, 1993 and successfully tested. We are currently in touch with Oakridge National Lab for replacement metal to enable the actual experiments to proceed.

B. Status of Research

Currently, the experimental set-up has been completed, tested and operational. We have been running experiments with the little lithium lead salvaged during the spill in scoping tests while awaiting for a new supply of uncontaminated metal. A request for a 3 months no-cost extension has been submitted to enable the research to be completed.

VIII. CONCLUSIONS

This status report has given the details of the work done up to March 31, 1993. While this is not the final report, it is clear that all the work involved in the design, assembly and testing of the experimental system have been completed. The three main programs for data acquisition and control, data analysis and liquid metal transport reaction models have been completed, tested and ready to be used for the research. The Oak Ridge National Lab is in the process of sending the needed supply of the lithium-lead metal. As a result of the delays, a request for a no cost extension to August 31, 1993 has been submitted. This will allow ample time for all the experiments to be run and analyzed and the final report completed.

REFERENCES

- [1] Herzog, J.P., "Lithium-Lead/Water Interactions: Experiments and Analysis", Ph.D. Dissertation, Fusion Technology Institute, Nuclear Engineering and Engineering Physics Department, University of Wisconsin - Madison.
- [2] Biney, P.O., Corradini, M.L. Lomperski, S. and Krueger, J., "A Mass Transport Model for Hydrogen Generation During Lithium-Lead/Water Interactions". Presented at the 13th Symposium of Fusion Energy, Knoxville, TN October 1989.
- [3] Collier, J., Convective Boiling and Condensation, McGraw-Hill 1972 pp 124-133.

NOMENCLATURE

A	Surface area
B₁	Arrhenius rate constant
C	Molar Concentration
D_o	Diffusion Coefficient proportionlity constant
D_{lm}	Diffusion Coefficient of liquid metal
E	Activation energy
H	Enthalpy
h	Molar enthalpy also as heat transfer coefficient
J	Molar flux
k	Thermal conductivity
M	Molecular weight
P	Pressure
g	Heat flux
R	Universal gas constant
S	Experimentally determined initial reaction rate
T	Absolute temperature
t	time
V	Velocity
Z	Principal coordinate direction
ρ	density

SUBSCRIPTS

pd	products
lm	liquid metal
cond	conduction
gen	generated
conv	convective
sat	saturated
Ar	Argon
fb	film boiling

APPENDIX A

DETERMINATION OF THEORETICAL HYDROGEN PRESSURE

The chemical reaction equation is given by

Here we have assumed there is enough water for the reaction to form lithium hydroxide. Assuming the lithium-lead in the lower chamber has a depth of 1 cm, then the valve is

$$\begin{aligned} V &= \pi D^2 L/4 \\ &= \pi (2.54)^2 (1)/4 \\ &= 5.07 \text{ cm}^3 = 5.07 \times 10^{-6} \text{ m}^3 \end{aligned}$$

The molecular weight of lithium lead is calculated using

$$M_{LiPb} = 0.17 (7) + 0.83 (207) = 173 \text{ kg}$$

The density of lithium lead is then given by

$$\begin{aligned} \rho_{Li_{17}Pb_{83}} &= (0.00688) (530) + (0.99312) (10444) \\ &= 10375.8 \text{ kg/m}^3 \end{aligned}$$

The mass of lithium lead is then given by

$$\begin{aligned} m_{Li_{17}Pb_{83}} &= \rho_{Li_{17}Pb_{83}} V \\ &= (10375.8 \text{ kg/m}^3) (5.07) (10^{-6} \text{ m}^3) \\ &= 0.0526 \text{ kg} \end{aligned}$$

Mass of H_2 Produced by reaction of .0526 kg of $Li_{17}Pb_{83}$ is given by

$$\begin{aligned} m &= 0.17 \times 0.0526/173 \\ &= 5.17 \times 10^{-5} \text{ kg} \end{aligned}$$

The hydrogen produced is assumed to behave as an ideal gas with gas constant

$$\begin{aligned} R_{H_2} &= R / M_{H_2} \\ &= 8.3143/2 \\ &= 4.1572 \text{ kJ/kg} \cdot \text{K} \end{aligned}$$

The hydrogen produced bubbles into the gas space in the upper chamber of inside diameter (1.5"). The height of the gas space above the water level in the upper chamber is varied for 5 cm to 25 cm. The volume of H_2 per centimeter height is given by

$$\begin{aligned} V_1 &= \pi D^2 (l)/4 \\ &= \pi (1.5 \times 2.54)^2 (1)/4 = 11.401 \text{ cc} \\ &= 11.401 \times 10^{-6} \text{ m}^3 \end{aligned}$$

The pressure of hydrogen is then given by

$$P_{H_2} = M_{H_2}RT/V$$

The maximum planned gas temperature is 100°C

Thus

$$\begin{aligned} P_{H_2} &= (0.0526) (4.1572) (373.15)/(11.401 \times 10^{-6}) \\ &= 19553 \text{ kPa} \end{aligned}$$

Table 2 gives the summary of the results of the calculations in this Appendix.

APPENDIX B

Listing of Data Analysis Program

```
C
C
C      PROGRAM
C
C      This program analyzes the experiment output. It reads in the
C      data from the countdown phase, the reaction phase and the equilibrium
c      phase from the datafiles COUNTD.DAT, REACTION.DAT and EQUIL.DAT
c      respectively.
c
c      It then calculates the initial pressure and temperatures in the upper
c      and lower chambers. It calculates the mass of argon in the initially in the lower
c      chamber at the start of the experiment. The program then
c      calculates the system pressure due to the water vapor
C      pressure and Argon pressure.
C
C
C      IMPLICIT DOUBLE PRECISION ( A-H, M-Z )
      IMPLICIT INTEGER ( I-L )
C      REAL *8 TLM,PLG,TLW,TLG
C
      DIMENSION TIME(500),UGT(500),TLG(500),UWT(500),TLW(500)
      DIMENSION TLM(500),UGP(500),PLG(500)
      COMMON/FACTOR/C(9)
      CHARACTER*15 FILENM1
      CHARACTER*15 FILENM2
      CHARACTER*15 FILENM3
      CHARACTER*15 FILEDP
      CHARACTER*1  RESPON
C
C
C***** PROGRAM CONSTANTS *****
C***** VUPFINAL IS THE FINAL GAS SPACE ABOVE LIQUID LEVEL AFTER EXPT.
C***** MEASURED FROM THE LIQUID LEVEL GAUGE AFTER THE EXPERIMENT.
C***** VDN = VOLUME OF THE LOWER CHAMBER BELOW THE LOWER FLANGE
C***** VDEAD1 = VOLUME OF DEAD SPACE IN VALVE
C***** VDEAD2 = VOLUME OF DEAD SPACE IN FLANGES AND THROUGH GASKET THICKNESS
C***** BELOW BALL OF CLOSED VALVE
C***** VUPTOTAL = TOTAL VOLUME OF THE UPPER CHAMBER
C***** FWL = FINAL WATER LEVEL IN INCHES
C
```

```

VDEAD1=22.86D0
VDEAD2= 25.17D0
VDN=90.09
VUPTOTAL = 1385.D0
ILEAK=0

```

```

C
C****CURVE FIT COEFFICIENTS FOR UPPER VOLUME AS FUNCTION OF WATER LEVEL
      ZO= -2.23958
      Z1= 28.9583

```

```

C
C*** EQUATION FOR UPPER VOLUME AS A FUNCTION OF FINAL WATER LEVEL

```

```

C
115 FORMAT(A)
      WRITE (IWRITE,145) FWL
145 FORMAT(10X,'The default final water level is ', 1P1G11.4, ' cm3'
3      / 10X, 'Do you want to change this (y/n)?' )
      READ (IREAD,115) RESPON
      IF ( ( RESPON .EQ. 'Y' ) .OR. ( RESPON .EQ. 'y' ) ) THEN
          WRITE (IWRITE,150)
150      FORMAT ( 10X, 'Enter the new final water level (in)')
          READ (IREAD,*) FWL
      ENDIF

```

```

      VUPFINAL=ZO+Z1*FWL

```

```

C
C **** INPUT THE LIQUID METAL MASS *****
C

```

```

      WRITE (6,125)
125  FORMAT ( 10X, 'Enter the liquid metal mass (gm)' )
      READ (IREAD,*) MASLIQ
      DLIQ = 9.
      VGASDN = VDN - MASLIQ / DLIQ + VDEAD2
      VGASUPI = VUPFINAL - (VGASDN + VDEAD1)
      VGASUPF = VUPFINAL
      VH2O=VUPTOTAL-VGASUPI
      RGAS = 83.14395
      MWH2O = 18.02

```

```

C***** SATURATION PRESSURE FORMULA FACTORS *****

```

```

      C(1) = -7.691234564
      C(2) = -26.08023696
      C(3) = -168.1706546
      C(4) = 64.23285504
      C(5) = -118.9646225
      C(6) = 4.167117320
      C(7) = 20.97506760
      C(8) = 1.D9
      C(9) = 6.

```

```

C***** LOGICAL UNIT NUMBERS *****

```

```

      IWRITE = 6
      IREAD = 6

```

```

IDATA1 = 9
IDATA2 = 10
IDATA3 = 11
IDELTP = 12
ITH2O = 13
ITLM = 14
IPRES = 15
ISYST = 16
IPCOMP = 17
IDOUT = 18
IPH2 = 19
IMH2 = 20
C***** OUTPUT FILES *****
  OPEN ( ITH2O, FILE = 'TH2O.DAT', STATUS = 'OLD' )
  OPEN ( ITLM, FILE = 'TLM.DAT', STATUS = 'OLD' )
  OPEN ( IPRES, FILE = 'PRES.DAT', STATUS = 'OLD' )
  OPEN ( ISYST, FILE = 'SYSTEM.TXT', STATUS = 'OLD' )
  OPEN ( IPH2, FILE = 'PH2.DAT', STATUS = 'OLD' )
  OPEN ( IMH2, FILE = 'MH2.DAT', STATUS = 'OLD' )
C***** INPUT SECTION *****
  FILENM1='A:COUNTD.DAT'
  FILENM2='A:REACTION.DAT'
  FILENM3='A:EQUIL.DAT'
  FILEDP='DELTP.DAT'
C  FILENM='P.DAT'
  OPEN ( IDATA1, FILE = FILENM1, STATUS = 'OLD' )
  OPEN ( IDATA2, FILE = FILENM2, STATUS = 'OLD' )
  OPEN ( IDATA3, FILE = FILENM3, STATUS = 'OLD' )
C
  IDTP = 1
  AREA = 5.0671
  STOT = MASLIQ / ( DLIQ * AREA )
C
  TEQ = 200.
  WRITE (IWRITE,165) TEQ
165  FORMAT ( 10X, 'The time at which the mass of hydrogen is aver',
1      'aged = ', 1P1G11.4, ' sec' / 10X, 'Do you want to ',
2      'change it (y/n)?' )
  READ (IREAD,115) RESPON
  IF ( ( RESPON .EQ. 'Y' ) .OR. ( RESPON .EQ. 'y' ) ) THEN
    WRITE (IWRITE,170)
170  FORMAT ( 10X, 'Enter the new value of TEQ ' )
    READ (IREAD,*) TEQ
  END IF
  IMHAVE = 0
  IFLAG = 0
  MH2SUM = 0.
C
  IPOUT = 0
  WRITE (IWRITE,175)
175  FORMAT ( 5X, 'Do you want the components of the pressure ',
1      'saved (y/n)?' )

```

```

      READ (IREAD,115) RESPON
      IF ( ( RESPON .EQ. 'Y' ) .OR. ( RESPON .EQ. 'y' ) ) THEN
        OPEN ( IPCOMP, FILE = 'PCOMP.DAT', STATUS = 'OLD' )
        IPOUT = 1
      END IF
C
200 CONTINUE
C
C*****
C  INITIAL EXPERIMENTAL DATA
C*****
C
C  In this section we read in the data for the countdown phase before the
C  valve is opened. We then calculate the average initial temperatures and
C  pressure in both the upper and lower chambers.
      I = 0
250 CONTINUE
C
      I = I + 1
C
      READ (IDATA1,*,END=251)TIME(I),UGT(I),TLG(I),UWT(I),TLW(I)
      1,TLM(I),UGP(I),PLG(I)
C
C***** CONVERT PRESSURE FROM VOLTS TO BARS
C
      UGP(I)=UGP(I)*20.D0*6.8948D-2
      PLG(I)=PLG(I)*20.D0*6.8948D-2
C
C  WRITE(*,*)TIME(I),UGT(I),TLG(I),UWT(I),TLW(I),TLM(I),UGP(I),PLG(I)

      GO TO 250
C
251 ICOUNT = I
      JCOUNT = ICOUNT - 1
C***** EVALUATE INITIAL LOWER GAS PRESSURE AND TEMPERATURE *****
C***** AND THE INITIAL UPPER GAS PRESSURE AND TEMPERATURE *****
C***** AND THE INITIAL LIQUID METAL TEMPERATURE *****
C ***** USING DATA FROM THE COUNTDOWN PHASE *****
      L = 0
      SUMPLG=0.
      SUMTLG=0.
      SUMUGP = 0.
      SUMUGT = 0.
      SUMUWT = 0.
      SUMTLM = 0.
C
C EVALUATE SUMS
C
      DO 300 I = 1, JCOUNT
        SUMPLG = SUMPLG + PLG(I)
        SUMTLG = SUMTLG + TLG(I)
C

```



```

SUMUGP = SUMUGP + UGP(I)
SUMUGT = SUMUGT + UGT(I)
SUMUWT = SUMUWT + UWT(I)
SUMTLM = SUMTLM + TLM(I)
C
300 CONTINUE
C
C EVALUATE MEAN VALUES
C
L=JCOUNT
PDN = SUMPLG / DBLE( L )
TDN = SUMTLG / DBLE( L )
PUP = SUMUGP / DBLE( L )
TUP = SUMUGT / DBLE( L )
TH2OI = SUMUWT / DBLE( L )
TLMI = SUMTLM / DBLE( L )
C***** EVALUATE STANDARD DEVIATIONS OF AVERAGED QUANTITIES *****
SDVPDN = 0.
SDVTDN = 0.
C
SDVPUP = 0.
SDVTUP = 0.
SDVTH2 = 0.
SDVTLM = 0.
C
DO 350 I = 1,JCOUNT
SDVPDN = SDVPDN + ( PDN - PLG(I) ) * ( PDN - PLG(I) )
SDVTDN = SDVTDN + ( TDN - TLG(I) ) * ( TDN - TLG(I) )
C
SDVPUP = SDVPUP + ( PUP - UGP(I) ) * ( PUP - UGP(I) )
SDVTUP = SDVTUP + ( TUP - UGT(I) ) * ( TUP - UGT(I) )
SDVTH2 = SDVTH2 + ( TH2OI - UWT(I) ) * ( TH2OI - UWT(I) )
SDVTLM = SDVTLM + ( TLMI - TLM(I) ) * ( TLMI - TLM(I) )

350 CONTINUE
SDVPDN = DSQRT( SDVPDN / DBLE( JCOUNT ) )
SDVTDN = DSQRT( SDVTDN / DBLE( JCOUNT ) )
C
SDVPUP = DSQRT( SDVPUP / DBLE( JCOUNT ) )
SDVTUP = DSQRT( SDVTUP / DBLE( JCOUNT ) )
SDVTH2 = DSQRT( SDVTH2 / DBLE( JCOUNT ) )
SDVTLM = DSQRT( SDVTLM / DBLE( JCOUNT ) )
C
C***** EVALUATE INITIAL WATER VAPOR PRESSURE *****
TUPK = TUP + 273.15
TDNK = TDN + 273.15
C
CALL PSAT (TUPK, PVAPI )
C
C***** EVALUATE INITIAL ARGON MASS IN SOLUTION *****
C

```

```

C   First we evaluate the initial Ar solubility and water density.
TSATK = 373.998 * ( PUP**(.07144015) )
TSAT = TSATK - 273.15
XAR = 3.7D-7 * ( TSAT - TH2OI ) * ( PUP - PVAPI ) / 1.013
IF(XAR .LT.0.0) XAR=0.0
DENH2O = 1.0098 - 4.86871D-4 * TH2OI

C
MARSOL = XAR * VH2O * DENH2O / MWH2O

C
MARUP = ( PUP - PVAPI ) * VGASUPI / (RGAS*TUPK)
MARDOWN = PDN * VGASDN / (RGAS*TDNK )
MAR = MARUP + MARDOWN + MARSOL
C***** OUTPUT INITIAL VALUES *****
C
C   First we output problem parameters to the screen.
WRITE (IWRITE,400) FILENM1,FILENM2,FILENM3
WRITE (ISYST,401) FILENM1,FILENM2,FILENM3
400  FORMAT ( /// 5X, 'For the system pressure files ', 3A12 / )
401  FORMAT ( 5X, 'For the system pressure files ', 3A12 / )
C
IF ( IDTP .EQ. 0 ) THEN
C   WRITE (IWRITE,405) FILEDP
C   WRITE (ISYST,405) FILEDP
405  FORMAT ( 5X, 'And the DELTAP pressure file ', A12 / )
END IF
C
WRITE (IWRITE,410) TLMI, SDVTLM, TH2OI, SDVTH2, MASLIQ, STOT
WRITE (ISYST,410) TLMI, SDVTLM, TH2OI, SDVTH2, MASLIQ, STOT
410  FORMAT( 5X, 'The experimental parameters are: ' /
1      5X, 'initial liquid metal temp. = ', 1P1G11.4, ' +/- ',
2      1P1G11.4, ' C ' /
3      5X, 'initial water temperature = ', 1P1G11.4, ' +/- ',
4      1P1G11.4, ' C ' /
5      5X, 'liquid metal mass = ', 1P1G11.4, ' gm ' /
6      5X, 'and the metal has a depth = ', 1P1G11.4, ' cm ' /)
C
WRITE (IWRITE,415) TUP, SDVTUP, VGASUPI, TDN, VGASDN
WRITE (ISYST,415) TUP, SDVTUP, VGASUPI, TDN, VGASDN
415  FORMAT(10X, 'with initial upper Ar temp = ', 1P1G11.4, ' +/- ',
1      1P1G11.4, ' C ' /
2      10X, 'the upper gas layer volume = ', 1P1G11.4, ' cm3 ' /
3      10X, 'with initial lower Ar temp = ', 1P1G11.4, ' C ' /
4      10X, 'the lower gas layer volume = ', 1P1G11.4, ' cm3 ' /)
C
WRITE (IWRITE,420) PUP, SDVPUP, PDN, SDVPDN
WRITE (ISYST,420) PUP, SDVPUP, PDN, SDVPDN
420  FORMAT(10X, 'the initial upper pressure = ', 1P1G11.4, ' +/- ',
2      1P1G11.4, ' bar ' /
3      10X, 'the initial lower pressure = ', 1P1G11.4, ' +/- ',
4      1P1G11.4, ' bar ' //)
C
WRITE (IWRITE,425) MAR, MARUP, MARDOWN, MARSOL

```

```

WRITE (ISYST,425) MAR, MARUP, MARDOWN, MARSOL
425  FORMAT(10X, 'the total argon mass      = ', 1P1G11.4, ' mole'/
1      10X, 'the upper chamb argon mass= ', 1P1G11.4, ' mole'/
1      10X, 'the lower chamb argon mass = ', 1P1G11.4, ' mole'/
1      10X, 'the argon mass in solution = ', 1P1G11.4, ' mole'/)

C
C  Now to output the initial values to the data files.
PAR = PUP - PVAPI
MHLEAK = 0.
IF ( IPOUT .EQ. 1 ) THEN
    WRITE (IPCOMP,998) TIME(JCOUNT), PAR, PVAPI, MARSOL, MHLEAK
END IF
PH2 = 0.
MH2TOT = 0.
C  WRITE (IPH2,999) TIME(JCOUNT), PH2
C  WRITE (IMH2,999) TIME(JCOUNT), MH2TOT
C
TIME0 = 0.0
C
C *****
C      MAIN CALCULATIONAL LOOP
C  THIS SECTION READS IN THE DATA FOR THE REACTION PHASE AND CALCULATES
C  THE MASS OF HYDROGEN GENERATED DURING THE PHASE AS A FUNCTION OF
C  TIME. CORRECTION FOR LEAKAGE IS MADE USING LEAKAGE INFORMATION
C  FROM A SEPARATE FILE
C *****
C *****
C      INNER LOOP
C *****
C
C ***** DATA INPUT SECTION *****
C
K=0

2000 CONTINUE
C
K=K+1
READ (IDATA2,*,END=2500)TIME1,TGAS1,TLG1,TH2O1,TLW1,TLM1
1 ,PRES1,PLG1
C
C ***** CONVERT PRESSURE FROM VOLTS TO BARS *****
C
PRES1=PRES1*20.D0*6.8948D-2
PLG1 =PLG1*20.D0*6.8948D-2

C  READ (IDATA,*,END=2500) TIME1, PRES1, TGAS1, TH2O1, TLM1
C
C
C ***** EVALUATE PVAP(TGAS1) *****
TGASK = TGAS1 + 273.15

```

```

C      CALL PSAT (TGASK,PVAP )
C
C***** EVALUATE ARGON MASS IN SOLUTION *****
C
C      This is the argon that stays in solution as it
c      bubbles to the upper chamber.
C
      TSATK = 373.998 * ( PRES1**(.07144015) )
      TSAT = TSATK - 273.15
      XAR = 3.7D-7 * ( TSAT - TH2O1 ) * PAR / 1.013
      DENH2O = 1.0098 - 4.86871D-4 * TH2O1
      IF(K.EQ.1) THEN
      XAR=0.D0
      ENDIF
C
      MARSOL = XAR * VH2O * DENH2O / MWH2O
C***** EVALUATE HYDROGEN MASS IN SOLUTION *****
C
C      This is the hydrogen in solution.
      XH2 = 1.7D-7 * ( TSAT - TH2O1 ) * PH2 / 1.013
      MH2SOL = XH2 * VH2O * DENH2O / MWH2O
C***** EVALUATE NEW ARGON MASS *****
C
C      Here we calculate the new argon mass, which is lower due to leak-
C age. But first we must estimate the leakage rate at the system pres-
C sure.
C ***** SET CONSTANTS TO ZERO *****
      A1=0.0
      B1=0.0
      C1=0.0
      IF(ILEAK .GT. 0 ) THEN
      DPDT = A1 + B1 * PRES1 + C1 * PRES1 * PRES1
      IF ( DPDT .LT. 0. ) DPDT = 0.
C
      MAR = MAR - ( DPDT * ( PAR / ( PAR + PH2+PVAP ) ) * ( TIME1
1          - TIME0 ) * VGASUPF / ( RGAS * TGASK ) )
C
C***** EVALUATE TOTAL HYDROGEN MASS TO HAVE LEAKED FROM SYSTEM *****
      MHLEAK = MHLEAK + ( DPDT * ( PH2 / ( PAR + PH2+PVAP ) ) *
1          ( TIME1 - TIME0 ) * VGASUPF / ( RGAS * TGASK ) )
      ENDIF
C
C***** UPDATE ARGON PRESSURE *****
      PAR = ( MAR - MARSOL ) * RGAS * TGASK / VGASUPF
C
C***** EVALUATE HYDROGEN PRESSURE *****
      PH2 = PRES1 - PAR - PVAP
C
C      PH2 = PRES1 - PAR - PVAP - DELPC - ( MH2SOL * RGAS * TGASK
C 1          / VGASUPF )

```

```

C
C***** EVALUATE TOTAL HYDROGEN MASS *****
      MH2TOT = ( PH2 * VGASUPF / ( RGAS * TGASK ) ) + MH2SOL +
1      MHLEAK

```

```

C
C
C***** OUTPUT NON-EQUILIBRIUM DATA *****
      WRITE (ITH2O,997) TIME1, TH2O1, TGAS1
      WRITE (ITLM,999) TIME1, TLM1
      WRITE (IPRES,999) TIME1, PRES1
      WRITE (IPH2,999) TIME1, PH2
      WRITE (IMH2,999) TIME1, MH2TOT
      IF ( IDTP .EQ. 0 ) THEN
C      WRITE (IDOUT,999) TIME1, DELPC
      END IF
      IF ( IPOUT .EQ. 1 ) THEN
      WRITE (IPCOMP,998) TIME1, PAR, PVAP, MARSOL, MHLEAK
      END IF

```

```

C      TIME0 = TIME1
C***** END OF INNER LOOP *****
C
      GO TO 2000
C
C***** END OF DELTAP COMPARISON LOOP *****
C
      2500 CONTINUE
C

```

```

C
C*****
C      MAIN CALCULATIONAL LOOP FOR EQUILIBRIUM PHASE
C      THIS SECTION READS IN THE DATA FOR THE EQUILIBRIUM PHASE AND CALCULATES
C      THE MASS OF HYDROGEN INVENTORY DURING THE PHASE AS A FUNCTION OF
C      TIME. CORRECTION FOR LEAKAGE IS MADE USING LEAKAGE INFORMATION
C      FROM A SEPARATE FILE
C*****
C*****
C      INNER LOOP
C*****
C
C***** DATA INPUT SECTION *****
C
      TIME0=0.0
      3000 CONTINUE
C
      READ (IDATA3,*,END=3500)TIME1,TGAS1,TLG1,TH2O1,TLW1,TLM1
      1 ,PRES1,PLG1
C

```

```

C ***** CONVERT PRESSURE FROM VOLTS TO BARS *****
C
  PRES1 = PRES1 * 20.D0 * 6.8948D-2
  PLG1 = PLG1 * 20.D0 * 6.8948D-2

C  WRITE(ISYST,*)TIME,TGAS1,TLG1
C  READ (IDATA,*,END=2500) TIME1, PRES1, TGAS1, TH2O1, TLM1
C
C
C ***** EVALUATE PVAP(TGAS1) *****
  TGASK = TGAS1 + 273.15
C
  CALL PSAT (
    1  TGASK,
    2
    3  PVAP )
C ***** EVALUATE ARGON MASS IN SOLUTION *****
C
C  This is the argon that comes out of solution due to the rising
C  water temperature.
  VH2O = VUPTOTAL - VUPFINAL
  TSATK = 373.998 * ( PRES1 ** (.07144015) )
  TSAT = TSATK - 273.15
  XAR = 3.7D-7 * ( TSAT - TH2O1 ) * PAR / 1.013
  DENH2O = 1.0098 - 4.86871D-4 * TH2O1
C
  MARSOL = XAR * VH2O * DENH2O / MWH2O
C ***** EVALUATE HYDROGEN MASS IN SOLUTION *****
C
C  This is the hydrogen in solution.
  XH2 = 1.7D-7 * ( TSAT - TH2O1 ) * PH2 / 1.013
  MH2SOL = XH2 * VH2O * DENH2O / MWH2O
C ***** EVALUATE NEW ARGON MASS *****
C
C  Here we calculate the new argon mass, which is lower due to leak-
C  age. But first we must estimate the leakage rate at the system pres-
C  sure.
  DPDT = A1 + B1 * PRES1 + C1 * PRES1 * PRES1
  IF ( DPDT .LT. 0. ) DPDT = 0.
C
  MAR = MAR - ( DPDT * ( PAR / ( PAR + PH2 ) ) * ( TIME1
    1  - TIME0 ) * VGASUPF / ( RGAS * TGASK ) )
C
C ***** EVALUATE TOTAL HYDROGEN MASS TO HAVE LEAKED FROM SYSTEM *****
C  MHLEAK = MHLEAK + ( DPDT * ( PH2 / ( PAR + PH2 ) ) *
    1  ( TIME1 - TIME0 ) * VGASUPF / ( RGAS * TGASK ) )
C
C ***** UPDATE ARGON PRESSURE *****
  PAR = ( MAR - MARSOL ) * RGAS * TGASK / VGASUPF
C
C ***** EVALUATE HYDROGEN PRESSURE *****

```

```

C
  PH2 = PRES1 - PAR - PVAP
  WRITE(*,*)"PRES1, PAR, PVAP, PH2",PRES1,PAR,PVAP,PH2
C
C  PH2 = PRES1 - PAR - PVAP - DELPC - ( MH2SOL * RGAS * TGASK
C  1    / VGASUPF )
C
C***** EVALUATE TOTAL HYDROGEN MASS *****
  MH2TOT = ( PH2 * VGASUPF / ( RGAS * TGASK ) ) + MH2SOL +
  1    MHLEAK
C  WRITE(*,*)"MADE IT THIS FAR"
C  STOP
C
C***** EVALUATE AVERAGE HYDROGEN MASS AT TEQ *****
C***** THIS IS DONE BY AVERAGING THE TOTAL MASS BETWEEN *****
C***** THE TIME INTERVAL TEQ-DTEQ AND TEQ
C  DTEQ=2.0
C  IF ( ( TIME1 .GE. ( TEQ - DTEQ ) ) .AND. ( TIME1 .LE.
C  1    ( TEQ + DTEQ ) ) ) THEN
    IMHAVE = IMHAVE + 1
    MH2SUM = MH2SUM + MH2TOT
    WRITE(ISYST,*)TIME1,MH2SUM
  C  ELSE
  C  IF ( ( TIME1 .GT. ( TEQ ) ) .AND. (IFLAG .EQ.
C  1    0 ) ) THEN
    IFLAG = 1
    MH2AVE = MH2SUM / DBLE( IMHAVE )
  C
  C  WRITE (IWRITE,450) TEQ, MH2AVE
  C  WRITE (ISYST,450) TEQ, MH2AVE
C 450  FORMAT(10X, 'at time              = ', 1P1G11.4,
C  1    ' sec ' /
C  2    5X, 'the total mass of hydrogen = ', 1P1G11.4,
C  3    ' mole' )
C
C  END IF
C  END IF
C
C***** OUTPUT DATA FROM NON-EQUILIBRIUM TO EQUILIBRIUM STATE *****

  WRITE (ITH2O,997) TIME1, Th2O1, TGAS1
  WRITE (ITLM,999) TIME1, TLM1
  WRITE (IPRES,999) TIME1, PRES1
  WRITE (IPH2,999) TIME1, PH2
  WRITE (IMH2,999) TIME1, MH2TOT
  IF ( IDTP .EQ. 0 ) THEN
C  WRITE (IDOUT,999) TIME1, DELPC
  END IF
  IF ( IPOUT .EQ. 1 ) THEN
    WRITE (IPCOMP,998) TIME1, PAR, PVAP, MARSOL, MHLEAK
  END IF
C

```

```

      TIME0 = TIME1
C***** END OF INNER LOOP *****
C
      GO TO 3000
C
C***** END OF DELTAP COMPARISON LOOP *****
C
3500 CONTINUE
C
      MH2AVE = MH2SUM / DBLE( IMHAVE )
C
      WRITE (IWRITE,450) TIME1, MH2AVE,MH2SOL,MHLEAK
      WRITE (ISYST,450) TIME1, MH2AVE,MH2SOL,MHLEAK
450  FORMAT(10X, 'at time              = ', 1P1G11.4,
1      ' sec ' /
2      5X, 'the total mass of hydrogen      = ', 1P1G11.4,
3      ' mole' /
2      5X, 'the total mass of hydrogen IN SOL = ', 1P1G11.4,
3      ' mole' /
2      5X, 'the total mass of hydrogen LEAKEED = ', 1P1G11.4,
3      ' mole' )
C
C
C*****
C      END OF PROGRAM
C*****
C
997  FORMAT ( 3(2X,1P1G11.4) )
998  FORMAT ( 5(2X,1P1G11.4) )
999  FORMAT ( 2(2X,1P1G11.4) )
C
C      That's all folks!
C      END
C
C*****
C
C*****
C
SUBROUTINE PSAT ( T,P )
C
C      This subroutine evaluates the the saturation pressure of water,
C      in Bar, given the water temperature in Kelvin. This subroutine uses
C      a formula for the saturation line given in the ASME steam tables.
C
      IMPLICIT DOUBLE PRECISION ( A-H, M-Z )
      IMPLICIT INTEGER ( I-L )
C
      COMMON / FACTOR / C(9)
C
C***** REDUCED TEMPERATURE *****

```



```

      THETA = T / 647.3
C***** FORMULA FACTORS *****
      THETA1 = 1. - THETA
C
      NUMER1 = 0.
      DO 100 I = 1, 5
          NUMER1 = C(I) * ( THETA1 ** DBLE(I) ) + NUMER1
100 CONTINUE
C
      DENOM1 = 1. + C(6) * THETA1 + C(7) * THETA1 * THETA1
C
      DENOM2 = C(8) * THETA1 * THETA1 + C(9)
C***** REDUCED PRESSURE *****
      BETA = DEXP( NUMER1 / ( DENOM1 * THETA ) - THETA1 / DENOM2 )
C***** SATURATION PRESSURE *****
      P = 221.2 * BETA
C
C    That's all folks!
      RETURN
      END

```

APPENDIX C

Listing of Data Acquisition and Control Program

```

10 REM *****
20 REM *   DATA ACQUISITION AND CONTROL PROGRAM FOR LITHIUM-LEAD/WATER   *
30 REM *           REACTION EXPERIMENTAL RESEARCH                       *
40 REM *           DEVELOPED BY                                           *
50 REM *           DR. PAUL ORLEANS BINEY                               *
60 REM *           JANUARY 1993                                           *
70 REM *****
80 REM
90 REM *****

100 REM * SET RELAY NUMBERS FOR DIGITAL RELAY CONTROLLED EQUIPMENT      *
110 REM * ALL RELAYS ARE ON PCM3, WHICH IS THE 16 CHANNEL AC/DC POWER    *
120 REM * CONTROL BOARD.                                                 *
130 REM *                                                                 *

140 REM * RELAY #  IONAME  DESCRIPTION OF EQUIPMENT ATTACHED          *
144 REM *  0      UWH     UPPER WATER HEATER                        *
150 REM *  1      LMH     LOWER LIQUID METAL HEATER                *
160 REM *  2      CWV     COOLING WATER SOLENOID VALVE              *
170 REM *  3      AIV     ARGON INLET SOLENOID VALVE                *
180 REM *  4      AOV     ARGON OUTLET SOLENOID VALVE              *
190 REM *  5      USV     UPPER SOLENOID VALVE (FOR PURGING SYSTEM)*
200 REM *  6      MBV     MAIN BALL VALVE ACTUATOR
210 REM *
220 REM *

230 REM * THERMOCOUPLES DESCRIPTION AND IONAMES ASSIGNED

240 REM * ALL THERMOCOUPLES ARE ON AIM7 BOARD
250 REM * IONAME: IS THE INPUT/OUTPUT NAME ASSIGNED IN CONFIGURATION FILE
260 REM * TYPE : IS THE THERMOCOUPLE TYPE
270 REM * GAIN : IS THE PRODUCT OF LOCAL GAIN (100) AND GLOBAL GAIN
280 REM * WTMAX : IS THE ANTICIPATED MAXIMUM WORKING TEMPERATURE
290 REM * VMAX : IS THE MAXIMUM OUTPUT VOLTAGE FROM TC AFTER GAIN IS
300 REM *          APPLIED. THE VOLTAGE CORRESPONDS TO THAT AT WTMAX
310 REM *          AIM7

320 REM *   DESCRIPTION   IONAME CHANNEL TYPE GAIN WTMAX  VMAX
330 REM *   -----
340 REM * UPPER GAS TEMP.   UGT   1    E 1000 150C 9.8680V
350 REM * LOWER GAS TEMP.   LGT   4    K  200 1000C 8.2536V
360 REM * UPPER WATER TEMP. UWT   2    K 1000 150C 6.1370V
370 REM * LOWER WATER TEMP. LWT   3    E 1000 150C 9.8680V
380 REM * UPPER FURNACE TEMP. UFT  6    K  200 1200C 9.7656V
390 REM * LOWER FURNACE TEMP. LFT  7    K  200 1200C 9.7656V

```

```

400 REM * LIQUID METAL TEMP. LMT 5 K 200 1000C 8.2536V
410 REM *

420 REM * PRESSURE TRANSDUCER DESCRIPTION AND IONAMES ASSIGNED IN
430 REM * CONFIGURATION FILE
440 REM * AMM2
450 REM * DESCRIPTION IONAME CHANNEL TYPE GAIN WPMAX VMAX
460 REM * -----
470 REM * UPPER GAS PRESSURE UGP 1 ABS 2 100 PSI 10.00V
475 REM * LOWER GAS PRESSURE LGP 2 ABS 2 100 PSI 10.00V
477 REM * *****

479 REM *****
481 REM * THE EXPERIMENT HAS FIVE MAJOR PHASES. THE DATA ACQUISITION *
483 REM * AND CONTROL SYSTEM HAS BEEN SET UP TO AUTOMATICALLY CONTROL *
485 REM * ALL EVENTS AND GO THROUGH ALL FIVE PHASES WITHOUT OPERATOR *
487 REM * INTERFACE. THE FOLLOWING ARE THE MAJOR VARIABLES THAT CAN *
489 REM * BE CHANGED IN THE PROGRAM IF NECESSARY BEFORE RUNNING *
491 REM * THE PROGRAM. *
492 REM * BACKGROUND BACKGROUND NUMBER NUMBER *
493 REM * PHASE DESCRIPTION INTERVAL INTERVAL OF DATA OF DATA*
494 REM * VARIABLE VALUE VARIABLE VALUE *
495 REM * 1 UPPER WATER HEATING BINT1% 200 NDP1! 90. *
496 REM * 2 LIQUID METAL HEATING BINT2% 200 NDP2! 120. *
497 REM * 3 COUNTDOWN TO REACTION BINT3% 10 NDP3! 60. *
498 REM * 4 CHEMICAL REACTION BINT4% 2 NDP4! 100. *
499 REM * 5 EQUILIBRIUM PERIOD BINT5% 200 NDP5! 90. *
500 REM * *
501 REM *****
502 REM TIME INTERVAL BETWEEN DATA=BINT*IR*0.001 SECONDS

503 BINT1% = 200: BINT2% = 200: BINT3% = 10: BINT4% = 2: BINT5% = 200 :BINTC% = 20
504 NDP1! = 110!: NDP2! = 140!: NDP3! = 80!: NDP4! = 120!: NDP5! = 110!: NEPC! = 1800!
505 DTSV=1 :BOILT=5.: RCDT=60. :DTM=10.0 :EQT=60. : TSAT=30.

506 REM * NRT IS THE NUMBER OF DATA POINTS CORRESPONDING TO REACTION TIME
510 REM * NRT=REACTION TIME/INTERRUPT PERIOD
520 REM *
530 REM * DEFINE DIGITAL ON/OFF VARIABLES
540 REM *
542 IR% = 100

544 DTUH = 25: UWSETPT = 30.: LMTSETPT = 100.: ITESTNO = 1: REACTIME% = 10
546 UWDT=10 : LMDT=10 :UFDT=50 :LFDT=50 :UFSETPT=20.0: LFSETPT=650.0
550 CALL KDINIT

600 DIM ONN%(1): DIM OFFF%(1): DIM TIM%(8): DIM TM#(8):DIM CDOWN(100,8)
602 DIM REACTT(150,8): DIM EQUIL(150,8)
603 DIM UWH1A(1), UWH1B(1), UWH1C(1), UWH1D(1), UWH1E(1): PRINT LP!
604 DIM LGH1A(1), LGH1B(1), LGH1C(1), LGH1D(1)

610 ONN%(0) = 0: OFFF%(0) = 1: ST% = 10: LP! = 1!

```

```

612 REM CWV$="CWV" : USV$="USV" : UWH$="UWH" : AIV$="AIV" : AOV$="AOV"
613 REM USV$="USV": UGT$="UGT" : LGT$="LGT": LFT$="LFT" : LMT$="LMT"
618 REM CJUFT$="CJN,UFT" : CJLGT$="CJN,LGT" : CJLMT$="CJN,LMT"

620 UGP$ = "UGP"
622 CWV$ = "CWV": USV$ = "USV": UWH$ = "UWH": AIV$ = "AIV": AOV$ = "AOV"
623 USV$ = "USV": UGT$ = "UGT": LGT$ = "LGT": LFT$ = "LFT": LMT$ = "LMT"
624 LMH$ = "LMH": UFT$ = "UFT": LGP$ = "LGP": UWT$ = "UWT": LWT$ = "LWT"
626 MBV$ = "MBV": CJUWLWT$ = "CJN,UWT,LWT": CJUGT$ = "CJN,UGT"
628 CJUFT$ = "CJN,UFT": CJLGT$ = "CJN,LGT": CJLMT$ = "CJN,LMT": CJLFT$="CJN,LFT"
629 REM

630 REM CREATE DIGITAL ARRAYS FOR CONTROLLING VALVES AND HEATERS

640 REM
650 CALL ARMAKE'("CWVOPEN%",1.,CWV$)
660 CALL ARMAKE'("CWVCLOSE%",1.,CWV$)
670 CALL ARMAKE'("USVOPEN%",1.,USV$)
680 CALL ARMAKE'("USVCLOSE%",1.,USV$)
690 CALL ARMAKE'("UWHONN%",1.,UWH$)
700 CALL ARMAKE'("UWHOFF%",1.,UWH$)
702 CALL ARMAKE'("LMHONN%",1.,LMH$)
705 CALL ARMAKE'("LMHOFF%",1.,LMH$)
710 CALL ARMAKE'("AIVOPEN%",1.,AIV$)
720 CALL ARMAKE'("AIVCLOSE%",1.,AIV$)
730 CALL ARMAKE'("AOVOPEN%",1.,AOV$)
740 CALL ARMAKE'("AOVCLOSE%",1.,AOV$)
744 CALL ARMAKE'("MBVOPEN%",1.,MBV$)
746 CALL ARMAKE'("MBVCLOSE%",1.,MBV$)

750 REM
760 REM PUT DIGITAL VALUES OF ON/OFF INTO KDAC ARRAYS FOR DIGITAL CONNTROL
770 REM OF VALVES AND HEATERS.

780 REM
800 CALL ARPUT'("CWVOPEN%",1.,1.,CWV$,1,ONN%,0,"C.RAW.INT")
810 CALL ARPUT'("CWVCLOSE%",1.,1.,CWV$,1,OFF%,0,"C.RAW.INT")
820 CALL ARPUT'("USVOPEN%",1.,1.,USV$,1,ONN%,0,"C.RAW.INT")
825 CALL ARPUT'("USVCLOSE%",1.,1.,USV$,1,OFF%,0,"C.RAW.INT")
830 CALL ARPUT'("UWHONN%",1.,1.,UWH$,1,ONN%,0,"C.RAW.INT")
840 CALL ARPUT'("UWHOFF%",1.,1.,UWH$,1,OFF%,0,"C.RAW.INT")
845 CALL ARPUT'("AIVOPEN%",1.,1.,AIV$,1,ONN%,0,"C.RAW.INT")
850 CALL ARPUT'("AIVCLOSE%",1.,1.,AIV$,1,OFF%,0,"C.RAW.INT")
860 CALL ARPUT'("AOVOPEN%",1.,1.,AOV$,1,ONN%,0,"C.RAW.INT")
870 CALL ARPUT'("AOVCLOSE%",1.,1.,AOV$,1,OFF%,0,"C.RAW.INT")
880 CALL ARPUT'("MBVOPEN%",1.,1.,MBV$,1,ONN%,0,"C.RAW.INT")
890 CALL ARPUT'("MBVCLOSE%",1.,1.,MBV$,1,OFF%,0,"C.RAW.INT")
891 CALL ARPUT'("LMHONN%",1.,1.,LMH$,1,ONN%,0,"C.RAW.INT")
892 CALL ARPUT'("LMHOFF%",1.,1.,LMH$,1,OFF%,0,"C.RAW.INT")
900 REM

910 REM BEGIN DATA ACQUISITION AND CONTROL

```

```

920 REM
921 PRINT " *****"
922 PRINT " * THE ARGON INLET VALVE IS BEING CHECKED AND OPENED      *"
923 PRINT " * THEN AFTER 2 INTONS  THE ARGON OUTLET VALVE IS CHECKED/OPENED*"
924 PRINT " * THE COOLING WATER SOLENOID VALVE IS BEING CLOSED      *"
925 PRINT " *****"

930 TNH = 0: TIM%(0) = 1
932 CALL KDTIMER'(TIM%(), "NT", "")
940 CALL INTON'(IR%, "MIL")
943 'CALL BGWRITE'("AIVOPEN%", AIV$, 1, 1, "NT", "AIVSTAT")
945 CALL KDTIMERRD'(TM#())
947 IF TM#(0)*IR%/1000 < DTSV GOTO 945
948 'CALL BGWRITE'("AOVOPEN%", AOV$, 1, 1, "NT", "AOVSTAT")
949 'CALL BGWRITE'("CWVOPEN%", CWV$, 1, 1, "NT", "CWVSTAT")

952 PRINT
953 PRINT " *****"
954 PRINT " * THE UPPER WATER HEATER AND LIQUID METAL HEATER ARE      *"
955 PRINT " * BEING TURNED ON TO BEGIN THE INITIAL HEATING PHASE.      *"
956 PRINT " *****"

960 CALL BGWRITE'("UWHONN%", UWH$, 1, 1, "NT", "UWHSTAT")
961 CALL BGWRITE'("LMHONN%", LMH$, 1, 1, "NT", "LMHSTAT")

962 UHST% = 1
970 CALL BGSTATUS'("CWVSTAT", ST%)
971 REM CALL BGWRITE'("USVCLOSE%", USV$, 1, 1, "NT", "USVSTAT2")
974 REM CALL BGWRITE'("MBVOPEN%", MBV$, 1, 1, "NT", "MBVSTAT2")
975 REM TIM%(1) = 1
976 REM CALL KDTIMER'(TIM%(), "NT", "")
977 REM CALL KDTIMERRD'(TM#())
978 REM IF TM#(1)*IR%/1000 < 5 GOTO 977
979 REM CALL BGWRITE'("MBVCLOSE%", MBV$, 1, 1, "NT", "MBVSTAT3")
980 REM IF TM#(1) < 6 GOTO 977
981 REM REM CALL BGWRITE'("MBVOPEN%", MBV$, 1, 1, "NT", "MBVSTAT3")
982 IF ST% < > 0 GOTO 970
983 PRINT "STATUS OF CWV IS ", ST%
984 REM CALL INTOFF
990 REM

992 PRINT " *****"
994 PRINT " * ACQUIRING DATA FOR HEATING OF UPPER WATER ON BACKGROUND GO *"
995 print " * AND SIMULTANEOUS DATA ACQUISITION OF LIQUID METAL TEMP.  *"
996 PRINT " *****"

1000 CALL BGREAD'("UWHEAT1A", NDP1!, CJUWLWT$, BINT1%, "NONE", 1, "NT", "UWHDAT1A")
1010 CALL BGREAD'("UWHEAT1B", NDP1!, CJUGT$, BINT1%, "NONE", 1, "NT", "UWHDAT1B")
1020 CALL BGREAD'("UWHEAT1C", NDP1!, CJUFT$, BINT1%, "NONE", 1, "NT", "UWHDAT1C")
1030 CALL BGREAD'("UWHEAT1D", NDP1!, UGP$, BINT1%, "NONE", 1, "NT", "UWHDAT1D")

1031 REM ***** READ LOWER DATA *****

```

```

1032 CALL BGREAD("LDAT0A",NDP1!,CJLGT$,BINT2%,"NONE",1,"NT","LDATA0A")
1034 CALL BGREAD("LDAT0B",NDP1!,CJLMT$,BINT2%,"NONE",1,"NT","LDATA0B")
1035 CALL BGREAD("LDAT0D",NDP1!,CJLFT$,BINT2%,"NONE",1,"NT","LDATA0D")
1036 CALL BGREAD("LDAT0C",NDP1!,LGP$,BINT2%,"NONE",1,"NT","LDATA0C")

1040 REM CALL INTON'(20,"SEC")
1042 REM CALL BGGO'("NT","")
1043 REM
1060 TNH = TNH + 1

1102 REM
1110 REM ***** GET LASTPOINT READ AND EXTRACT VALUES USING ARGET *****
1120 REM

1125 LP!=0.
1128 LPO! = LP!
1130 CALL ARLASTP'("UWHEAT1A",LP!)
1131 IF ABS(LPO! - LP!) < .1 GOTO 1128
1135 CALL ARGET'("UWHEAT1A",LP!,LP!,UWT$,1,UWH1A!(), "C.THCU.K")
1140 CALL ARGET'("UWHEAT1A",LP!,LP!,LWT$,1,UWH1B!(), "C.THCU.E")
1150 CALL ARGET'("UWHEAT1B",LP!,LP!,UGT$,1,UWH1C!(), "C.THCU.E")
1160 CALL ARGET'("UWHEAT1C",LP!,LP!,UFT$,1,UWH1D!(), "C.THCU.K")
1170 CALL ARGET'("UWHEAT1D",LP!,LP!,UGP$,1,UWH1E!(), "C.VOLTS")

1171 REM ***** GET LOWER DATA *****
1172 CALL ARGET'("LDAT0A",LP!,LP!,LGT$,1,LGH1A!(), "C.THCU.K")
1173 CALL ARGET'("LDAT0B",LP!,LP!,LMT$,1,LGH1B!(), "C.THCU.K")
1174 CALL ARGET'("LDAT0D",LP!,LP!,LFT$,1,LGH1D!(), "C.THCU.K")
1175 CALL ARGET'("LDAT0C",LP!,LP!,LGP$,1,LGH1C!(), "C.VOLTS")

1176 REM ***** PRINT DATA *****
1177 'PRINT "UWT,LWT,UGT,UFT,LMT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0),
UWH1D(0),LGH1B!(0), UWH1E(0)
1178 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0):UWAVG=0.5*(UWT1+LWT1)
1179 UGP1=UWH1E!(0)*20*6.8948E-2: LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
1180 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0):UFT1=UWH1D!(0)
1181 PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
1182 PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

1183 IF( UFT1 > UFSETPT+UFDT AND UHST%=1) THEN GOSUB 4000
1184 IF( UFT1 < UFSETPT-UFDT AND UHST%=0) THEN GOSUB 5000
1185 IF( LFT1 > LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
1186 IF( LFT1 < LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

1187 'PRINT "UWT,LWT,UGT,UFT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0), UWH1D(0),
UWH1E(0)

1190 REM
1200 REM ***** PERFORM TEST TO DETERMINE IF BOILING HAS STARTED *****
1210 REM

1220 IF UWH1A(0) < TSAT! GOTO 1128

```

```

1222 REM IF UWH1E!(0) > 1.5 GOTO 1128

1230 REM
1240 PRINT " ***** BOILING HAS STARTED *****"
1250 PRINT " ***** A 5-MINUTE BOILING COUNTDOWN TIME IS IN PROGRESS *****"
1260 REM

1270 TIM%(1) = 1
1280 CALL KDTIMER'(TIM%(),"NT","")
1290 CALL KDTIMERRD'(TM#())
1291 LPO! = LP!
1292 CALL ARLASTP'("UWHEAT1A",LP!)
1293 IF ABS(LPO! - LP!) < .1 GOTO 1290

1294 CALL ARGET'("UWHEAT1A",LP!,LP!,UWT$,1,UWH1A!(), "C.THCU.K")
1295 CALL ARGET'("UWHEAT1A",LP!,LP!,LWT$,1,UWH1B!(), "C.THCU.E")
1296 CALL ARGET'("UWHEAT1B",LP!,LP!,UGT$,1,UWH1C!(), "C.THCU.E")
1297 CALL ARGET'("UWHEAT1C",LP!,LP!,UFT$,1,UWH1D!(), "C.THCU.K")
1298 CALL ARGET'("UWHEAT1D",LP!,LP!,UGP$,1,UWH1E!(), "C.VOLTS")

1299 REM ***** GET LOWER DATA *****
1301 CALL ARGET'("LDAT0A",LP!,LP!,LGT$,1,LGH1A!(), "C.THCU.K")
1302 CALL ARGET'("LDAT0B",LP!,LP!,LMT$,1,LGH1B!(), "C.THCU.K")
1303 CALL ARGET'("LDAT0D",LP!,LP!,LFT$,1,LGH1D!(), "C.THCU.K")
1304 CALL ARGET'("LDAT0C",LP!,LP!,LGP$,1,LGH1C!(), "C.VOLTS")

1305 REM ***** PRINT DATA *****
1306 'PRINT "UWT,LWT,UGT,UFT,LMT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0),
UWH1D(0),LGH1B!(0), UWH1E(0)

1307 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0): UWAVG=0.5*(UWT1+LWT1)
1308 UGP1=UWH1E!(0)*20*6.8948E-2 : LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
1309 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0): UFT1=UWH1D!(0)
1310 PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
1311 PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

1312 IF( UFT1>UFSETPT+UFDLT AND UHST%=1) THEN GOSUB 4000
1313 IF( UFT1<UFSETPT-UFDLT AND UHST%=0) THEN GOSUB 5000
1314 IF( LFT1>LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
1315 IF( LFT1<LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

1316 'PRINT "UWT,LWT,UGT,UFT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0), UWH1D(0),
UWH1E(0)
1317 IF TM#(1)*IR%/1000 < BOILT GOTO 1290

1318 REM
1319 PRINT " *** COUNTDOWN TIME IS OVER "
1320 PRINT " *** THE UPPER SOLENOID VALVE IS BEING CLOSED"
1330 PRINT " *** THE COOLING WATER VALVE IS BEING OPENED"
1332 REM

1360 'CALL BGWRITE'("USVCLOSE%",USV$,1,1,"NT","USVSTAT2")

```

```

1370 CALL BGWRITE('CWVOPEN%',CWV$,1,1,"NT","CWVSTAT2")
1371 CALL BGWRITE('UWHOFFF%',UWH$,1,1,"NT","UWHSTAT2")
1372 TIM%(2) = 1 :REM MAY HAVE TO INSERT COOLING PERIOD FOR UPPER WATER
1374 CALL KDTIMER'(TIM%(),"NT","")
1380 CALL KDTIMERRD'(TM#())
1390 IF TM#(2)*IR%/1000 < DTSV GOTO 1380

1400 REM
1410 PRINT " ***** THE UPPER WATER HEATER IS BEING TURNED ON ***** "
1420 PRINT " ***** THE LIQUID METAL HEATER IS BEING TURNED ON***** "
1430 REM STOP

1440 CALL BGWRITE('UWHONN%',UWH$,1,1,"NT","UWHSTAT2")
1442 UHST%=1
1450 CALL BGWRITE('LMHONN%',LMH$,1,1,"NT","LMHSTAT2")
1452 LHST%=1
1460 'CALL BGHALT'('UWHDAT1A,UWHDAT1B,UWHDAT1C,UWHDAT1D',"NT","HALT1")

1470 REM
1480 PRINT " ***** NOW ACQUIRING DATA FOR LIQUID METAL HEATING "
1482 PRINT " ***** AND TEMPERATURE CONTROL
1485 PRINT " ***** PHASE ON BACKGROUND GO "
1490 REM ***** ACQUIRE UPPER DATA *****

1500 CALL BGREAD'('UWHEAT2A",NDP2!,CJUWLWT$,BINT2%,"NONE",1,"NT","UWHDAT2A")
1510 CALL BGREAD'('UWHEAT2B",NDP2!,CJUFT$,BINT2%,"NONE",1,"NT","UWHDAT2B")
1530 CALL BGREAD'('UWHEAT2C",NDP2!,CJUFT$,BINT2%,"NONE",1,"NT","UWHDAT2C")
1535 CALL BGREAD'('UWHEAT2D",NDP2!,UGP$,BINT2%,"NONE",1,"NT","UWHDAT2D")

1537 REM ***** ACQUIRE LOWER DATA *****
1540 CALL BGREAD'('LDAT1A",NDP2!,CJLGT$,BINT2%,"NONE",1,"NT","LDAT1A")
1544 CALL BGREAD'('LDAT1B",NDP2!,CJLMT$,BINT2%,"NONE",1,"NT","LDAT1B")
1545 CALL BGREAD'('LDAT1C",NDP2!,LGP$,BINT2%,"NONE",1,"NT","LDAT1C")
1548 CALL BGREAD'('LDAT1D",NDP2!,CJLFT$,BINT2%,"NONE",1,"NT","LDAT1D")
1550 REM CALL INTON'(20,"SEC")
1552 REM CALL BGGO'("NT","")

1560 REM
1570 REM ***** GET LASTPOINT READ AND EXTRACT VALUES USING ARGET *****
1580 REM

1592 LP! = 0!
1600 LPO! = LP!
1602 CALL ARLASTP'('LDAT1C",LP!)
1603 IF ABS(LPO! - LP!) < .1 GOTO 1600

1604 REM ***** GET UPPER DATA *****
1610 CALL ARGET'('UWHEAT2A",LP!,LP!,UWT$,1,UWH1A!(),"C.THCU.K")
1620 CALL ARGET'('UWHEAT2A",LP!,LP!,LWT$,1,UWH1B!(),"C.THCU.E")
1630 CALL ARGET'('UWHEAT2B",LP!,LP!,UGT$,1,UWH1C!(),"C.THCU.E")
1640 CALL ARGET'('UWHEAT2C",LP!,LP!,UFT$,1,UWH1D!(),"C.THCU.K")
1650 CALL ARGET'('UWHEAT2D",LP!,LP!,UGP$,1,UWH1E!(),"C.VOLTS")

```



```

1652 REM ***** GET LOWER DATA *****
1660 CALL ARGET('LDAT1A',LP!,LP!,LGT$,1,LGH1A!(), "C.THCU.K")
1670 CALL ARGET('LDAT1B',LP!,LP!,LMT$,1,LGH1B!(), "C.THCU.K")
1672 CALL ARGET('LDAT1D',LP!,LP!,LFT$,1,LGH1D!(), "C.THCU.K")
1673 CALL ARGET('LDAT1C',LP!,LP!,LGP$,1,LGH1C!(), "C.VOLTS")

1674 REM ***** PRINT DATA *****
1675 PRINT "UWT,LWT,UGT,UFT,LMT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0),
UWH1D(0),LGH1B!(0), UWH1E(0)
1676 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0): UWAVG=0.5*(UWT1+LWT1)
1677 UGP1=UWH1E!(0)*20*6.8948E-2: LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
1678 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0): UFT1=UWH1D!(0)
1679 PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
1680 PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

1681 IF( UFT1 > UFSETPT+UFDT AND UHST%=1) THEN GOSUB 4000
1682 IF( UFT1 < UFSETPT-UFDT AND UHST%=0) THEN GOSUB 5000
1683 IF( LFT1 > LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
1684 IF( LFT1 < LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

1722 IF LMT1 < (LMTSETPT - DTM) GOTO 1600

1724 REM
1726 PRINT " *****"
1728 PRINT " * THE LIQUID METAL HAS ALMOST REACHED ITS SET TEMPERATURE * "
1730 PRINT " * THE ARGON OUTLET VALVE IS BEING CLOSED * "
1732 PRINT " * THE ARGON INLET VALVE IS BEING CLOSED 1 SEC LATER * "
1734 PRINT " * A 60 SECONDS COUNTDOWN TIME IS BEING STARTED TO ACQUIRE * "
1736 PRINT " * DATA FOR CALCULATING THE LOWER ARGON MASS AND FOR * "
1738 PRINT " * ESTABLISHING INITIAL CONDITIONS * "
1740 PRINT " *****"

1742 REM
1744 REM *** CLOSE THE ARGON OUTLET VALVE AND THE INLET VALVE
1746 REM

1748 'CALL BGWRITE'('AOVCLOSE%',AOV$,1,1,"NT","AOVSTAT2")
1750 TIM%(3) = 1
1752 CALL KDTIMER'(TIM%(), "NT", "")
1754 CALL KDTIMERRD'(TM#())
1756 IF TM#(3)*IR%/1000 < DTSV GOTO 1754
1758 'CALL BGWRITE'('AIVCLOSE%',AIV$,1,1,"NT","AIVSTAT2")
1760 CALL BGHALT'('UWHDAT2A,UWHDAT2B,UWHDAT2C,UWHDAT2D',"NT","HALT1")
1762 CALL BGHALT'('LDATA1A,LDATA1B,LDATA1C',"NT","HALT2")

1764 PRINT
1766 PRINT " PROIGRAM IS NOW CONTROLLING THE UPPER WATER"
1768 PRINT " TEMPERATURE, AND THE LIQUID METAL TEMPERATURE"
1770 PRINT " UNTIL THEY ARE IN THE PROPER SETPOINTS"
1772 PRINT
1774 REM

```

```

1775 CALL BGREAD'("UWHEATCA",NDPC!,CJUWLWT$,BINT2%,"NONE",1,"NT","UWHDATCA")
1776 CALL BGREAD'("UWHEATCB",NDPC!,CJUGT$,BINT2%,"NONE",1,"NT","UWHDATCB")
1777 CALL BGREAD'("UWHEATCC",NDPC!,CJUFT$,BINT2%,"NONE",1,"NT","UWHDATCC")
1778 CALL BGREAD'("UWHEATCD",NDPC!,UGP$,BINT2%,"NONE",1,"NT","UWHDATCD")

1779 REM ***** ACQUIRE LOWER DATA *****
1780 CALL BGREAD'("LDATCA",NDPC!,CJLGT$,BINT2%,"NONE",1,"NT","LDATACA")
1781 CALL BGREAD'("LDATCB",NDPC!,CJLMT$,BINT2%,"NONE",1,"NT","LDATACB")
1782 CALL BGREAD'("LDATCC",NDPC!,LGP$,BINT2%,"NONE",1,"NT","LDATACC")
1783 CALL BGREAD'("LDATCD",NDPC!,CJLFT$,BINT2%,"NONE",1,"NT","LDATACD")
1784 REM CALL INTON'(20,"SEC")
1785 REM CALL BGGO'("NT","")

1860 REM
1870 REM ***** GET LASTPOINT READ AND EXTRACT VALUES USING ARGET *****
1880 REM

1882 REM DIM LGH1A(1), LGH1B(1), LGH1C(1)
1884 LP! = 0!
1886 LPO! = LP!
1888 CALL ARLASTP'("LDATCD",LP!)
1890 IF ABS(LPO! - LP!) < .1 GOTO 1886

1891 REM ***** GET UPPER DATA *****
1892 CALL ARGET'("UWHEATCA",LP!,LP!,UWT$,1,UWH1A!(), "C.THCU.K")
1893 CALL ARGET'("UWHEATCA",LP!,LP!,LWT$,1,UWH1B!(), "C.THCU.E")
1894 CALL ARGET'("UWHEATCB",LP!,LP!,UGT$,1,UWH1C!(), "C.THCU.E")
1895 CALL ARGET'("UWHEATCC",LP!,LP!,UFT$,1,UWH1D!(), "C.THCU.K")
1896 CALL ARGET'("UWHEATCD",LP!,LP!,UGP$,1,UWH1E!(), "C.VOLTS")

1897 REM ***** GET LOWER DATA *****
1898 CALL ARGET'("LDATCA",LP!,LP!,LGT$,1,LGH1A!(), "C.THCU.K")
1899 CALL ARGET'("LDATCB",LP!,LP!,LMT$,1,LGH1B!(), "C.THCU.K")
1900 CALL ARGET'("LDATCC",LP!,LP!,LGP$,1,LGH1C!(), "C.VOLTS")
1901 CALL ARGET'("LDATCD",LP!,LP!,LFT$,1,LGH1D!(), "C.THCU.K")

1902 REM ***** PRINT DATA *****
1903 PRINT "UWT,LWT,UGT,UFT,LMT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0),
UWH1D(0),LGH1B!(0), UWH1E(0)
1904 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0):UWAVG=0.5*(UWT1+LWT1)
1905 UGP1=UWH1E!(0)*20*6.8948E-2 : LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
1906 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0):UFT1=UWH1D!(0)
1907 PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
1908 PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

1909 IF( UFT1>UFSETPT+UFDI AND UHST%=1) THEN GOSUB 4000
1910 IF( UFT1<UFSETPT-UFDI AND UHST%=0) THEN GOSUB 5000
1911 IF( LFT1>LFSETPT+LFDI AND LHST%=1) THEN GOSUB 4500
1912 IF( LFT1<LFSETPT-LFDI AND LHST%=0) THEN GOSUB 5500

1913 REM ***** PERFORM MAJOR TEST TO BEGIN REACTION *****

```

```

1914 IF ABS(UWT1-LWT1)<UWDT AND ((UWAVG<UWSETPT+UWDT)) AND
(UWAVG>(UWSETPT-UWDT)) AND (LMT1<LMTSETPT+LMDT) AND (LMT1>LMTSETPT-LMDT)
GOTO 1929

```

```

1927 GOTO 1886

```

```

1929 PRINT " LIQUID METAL AND UPPER WATER ARE AT PRESET TEMPS"
1930 PRINT "UWT, LWT, LMT,UGP ARE ";UWT1;LWT1;LMT1;UGP1*20.0*6.8948E-2
2030 PRINT " ***** ACQUIRE DATA FOR COUNTDOWN *****"
2040 PRINT " ***** PHASE ON BACKGROUND GO *****"

```

```

2050 REM ***** ACQUIRE UPPER DATA *****
2052 TIM%(4) = 1
2054 CALL KDTIMER'(TIM%(), "NT", "")

```

```

2057 CALL BGREAD'("UWHEAT3A",NDP3!,CJUWLWT$,BINT3%,"NONE",1,"NT","UWHDAT3A")
2058 CALL BGREAD'("UWHEAT3B",NDP3!,CJUPT$,BINT3%,"NONE",1,"NT","UWHDAT3B")
2059 CALL BGREAD'("UWHEAT3C",NDP3!,CJUFT$,BINT3%,"NONE",1,"NT","UWHDAT3C")
2060 CALL BGREAD'("UWHEAT3D",NDP3,UGP$,BINT3%,"NONE",1,"NT","UWHDAT3D")

```

```

2061 REM ***** ACQUIRE LOWER DATA *****
2062 CALL BGREAD'("LDAT2A",NDP3!,CJLGT$,BINT3%,"NONE",1,"NT","LDATA2A")
2063 CALL BGREAD'("LDAT2B",NDP3!,CJLMT$,BINT3%,"NONE",1,"NT","LDATA2B")
2064 CALL BGREAD'("LDAT2C",NDP3!,LGP$,BINT3%,"NONE",1,"NT","LDATA2C")
2065 CALL BGREAD'("LDAT2D",NDP3!,CJLFT$,BINT3%,"NONE",1,"NT","LDATA2D")
2066 REM CALL INTON'(20,"SEC")
2067 REM CALL BGGO'("NT", "")

```

```

2144 LP3! = 0!
2145 LPO! = LP3!
2147 CALL ARLASTP'("LDAT2D",LP3!)
2149 IF ABS(LPO! - LP3!) < .1 GOTO 2145

```

```

2150 REM ***** GET UPPER DATA *****
2151 CALL KDTIMERRD'(TM#0)
2152 CALL ARGET'("UWHEAT3A",LP3!,LP3!,UWT$,1,UWH1A!0,"C.THCU.K")
2154 CALL ARGET'("UWHEAT3A",LP3!,LP3!,LWT$,1,UWH1B!0,"C.THCU.E")
2156 CALL ARGET'("UWHEAT3B",LP3!,LP3!,UGT$,1,UWH1C!0,"C.THCU.E")
2158 CALL ARGET'("UWHEAT3C",LP3!,LP3!,UFT$,1,UWH1D!0,"C.THCU.K")
2160 CALL ARGET'("UWHEAT3D",LP3!,LP3!,UGP$,1,UWH1E!0,"C.VOLTS")

```

```

2161 REM ***** GET LOWER DATA *****

```

```

2162 CALL ARGET'("LDAT2A",LP3!,LP3!,LGT$,1,LGH1A!0,"C.THCU.K")
2163 CALL ARGET'("LDAT2B",LP3!,LP3!,LMT$,1,LGH1B!0,"C.THCU.K")
2164 CALL ARGET'("LDAT2C",LP3!,LP3!,LGP$,1,LGH1C!0,"C.VOLTS")
2165 CALL ARGET'("LDAT2D",LP3!,LP3!,LFT$,1,LGH1D!0,"C.THCU.K")

```

```

2166 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0): UWAVG=0.5*(UWT1+LWT1)
2167 UGP1=UWH1E!(0)*20*6.8948E-2: LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)

```

```

2168 LGT1=LGHI1A!(0): LMT1=LGHI1B!(0): LFT1=LGHI1D!(0):UFT1=UWH1D!(0)
2169 'PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
2170 'PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

2171 IF( UFT1>UFSETPT+UFDT AND UHST%=1) THEN GOSUB 4000
2172 IF( UFT1<UFSETPT-UFDT AND UHST%=0) THEN GOSUB 5000
2173 IF( LFT1>LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
2174 IF( LFT1<LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

2175 TIME3=TM#(4)*IR%/1000
2176 CDOWN(LP3,1)=TIME3: CDOWN(LP3,2)=UWH1C!(0) :CDOWN(LP3,3)=LGHI1A!(0):
CDOWN(LP3,4)=UWH1A!(0)
2177 CDOWN(LP3,5)=UWH1B!(0): CDOWN(LP3,6)=LGHI1B!(0): CDOWN(LP3,7)=UWH1E!(0):
CDOWN(LP3,8)=LGHI1C!(0)

2178 REM PRINT DATA *****
2179 PRINT "TIME3,UWT,LWT,UGT,UFT,LGT,LMT,UGP,LGP ARE
",TIME3,UWH1A(0),UWH1B(0),UWH1C(0),UWH1D(0),LGHI1A!(0),LGHI1B!(0),UWH1E(0),LGHI1C!(0)
2180 REM
2181 REM

2182 'CALL KDTIMERRD'(TM#(0))
2190 IF TM#(4)*IR%/1000 < RCDT GOTO 2145

2200 REM
2210 PRINT : PRINT " *** COUNTDOWN TIME IS OVER*****"
2220 PRINT " *** THE MAIN BALL VALVE IS BEING OPENED *****"
2222 REM

2230 REM CALL BGWRITE'("MBVOPEN%",MBV$,1,1,"WGO","MBVSTAT1")
2240 CALL BGHALT'("UWHDAT3A,UWHDAT3B,UWHDAT3C,UWHDAT3D","WGO","HALT3")
2250 CALL BGHALT'("LDATA2A,LDATA2B,LDATA2C","WGO","HALT4")

2260 TIM%(5) = 1
2270 CALL KDTIMER'(TIM%(), "WGO", "")

2280 REM
2290 PRINT : PRINT " ***** ACQUIRING DATA FOR REACTION*****"
2300 PRINT " ***** PHASE ON BACKGROUND GO *****"
2302 REM
2310 REM ***** ACQUIRE UPPER DATA *****
2312 REM

2320 CALL BGREAD'("UWHEAT4A",NDP4!,CJUWLWT$,BINT4%,"NONE",1,"WGO","UWHDAT4A")
2330 CALL BGREAD'("UWHEAT4B",NDP4!,CJUWT$,BINT4%,"NONE",1,"WGO","UWHDAT4B")
2340 CALL BGREAD'("UWHEAT4C",NDP4!,CJUFT$,BINT4%,"NONE",1,"WGO","UWHDAT4C")
2350 CALL BGREAD'("UWHEAT4D",NDP4!,UGP$,BINT4%,"NONE",1,"WGO","UWHDAT4D")

2352 REM
2360 REM ***** ACQUIRE LOWER DATA *****
2361 REM

```

```

2362 CALL BGREAD'("LDAT3A",NDP4!,CJLGT$,BINT4%,"NONE",1,"WGO","LDATE3A")
2363 CALL BGREAD'("LDAT3B",NDP4!,CJLMT$,BINT4%,"NONE",1,"WGO","LDATE3B")
2364 CALL BGREAD'("LDAT3C",NDP4!,LGP$,BINT4%,"NONE",1,"WGO","LDATE3C")
2365 CALL BGREAD'("LDAT3D",NDP4!,CJLFT$,BINT4%,"NONE",1,"WGO","LDATE3D")
2366 REM CALL INTON'(100,"MIL")
2367 CALL BGGO'("NT","")
2368 REM

2369 LP4! = 0!
2370 LPO! = LP4!
2371 CALL ARLASTP'("LDAT3D",LP4!)
2372 IF ABS(LPO! - LP4!) < .1 GOTO 2370
2373 CALL KDTIMERRD'(TM#())

2374 REM ***** GET UPPER DATA *****
2376 CALL ARGET'("UWHEAT4A",LP4!,LP4!,UWT$,1,UWH1A!(), "C.THCU.K")
2378 CALL ARGET'("UWHEAT4A",LP4!,LP4!,LWT$,1,UWH1B!(), "C.THCU.E")
2380 CALL ARGET'("UWHEAT4B",LP4!,LP4!,UGT$,1,UWH1C!(), "C.THCU.E")
2381 'CALL ARGET'("UWHEAT4C",LP4!,LP4!,UFT$,1,UWH1D!(), "C.THCU.K")
2382 CALL ARGET'("UWHEAT4D",LP4!,LP4!,UGP$,1,UWH1E!(), "C.VOLTS")

2383 REM ***** GET LOWER DATA *****
2384 CALL ARGET'("LDAT3A",LP4!,LP4!,LGT$,1,LGH1A!(), "C.THCU.K")
2385 CALL ARGET'("LDAT3B",LP4!,LP4!,LMT$,1,LGH1B!(), "C.THCU.K")
2386 CALL ARGET'("LDAT3C",LP4!,LP4!,LGP$,1,LGH1C!(), "C.VOLTS")
2386 'CALL ARGET'("LDAT3D",LP4!,LP4!,LFT$,1,LGH1D!(), "C.THCU.K")

2387 TIME4=TM#(5)*IR%/1000

2388 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0): UWAVG=0.5*(UWT1+LWT1)
2389 UGP1=UWH1E!(0)*20*6.8948E-2 : LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
2390 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0): UFT1=UWH1D!(0)
2391 'PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
2392 'PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

2394 'IF( UFT1>UFSETPT+UFDT AND UHST%=1) THEN GOSUB 4000
2395 'IF( UFT1<UFSETPT-UFDT AND UHST%=0) THEN GOSUB 5000
2396 'IF( LFT1>LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
2397 'IF( LFT1<LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

2398 REACTT(LP4,1)=TIME4: REACTT(LP4,2)=UWH1C!(0) :REACTT(LP4,3)=LGH1A!(0):
REACTT(LP4,4)=UWH1A!(0)
2399 REACTT(LP4,5)=UWH1B!(0): REACTT(LP4,6)=LGH1B!(0): REACTT(LP4,7)=UWH1E!(0):
REACTT(LP4,8)=LGH1C!(0)

2400 REM PRINT DATA *****
2401 'PRINT "TIME4,UWT,LWT,UGT,UFT,LGT,LMT,UGP,LGP ARE
",TIME4,UWH1A(0),UWH1B(0),UWH1C(0),UWH1D(0),LGH1A!(0),LGH1B!(0),UWH1E(0),LGH1C!(0)
2402 REM

2430 'CALL KDTIMERRD'(TM#())
2440 IF TM#(5) * IR%/1000 < REACTIME% GOTO 2370

```

```

2450 REM
2460 PRINT " *** REACTION IS OVER "
2470 PRINT " *** THE MAIN BALL VALVE IS BEING CLOSED "
2480 PRINT " *** LIQUID METAL HEATER IS BEING TURNED OFF "
2490 PRINT " *** UPPER WATER HEATER IS BEING TURNED OFF "
2500 PRINT " *** ARGON OUTLET VALVE IS BEING OPENED 2 SEC AFTER BALL VALVE CLOSURE"
2510 PRINT " *** SETTING-UP TO ACQUIRE UPPER CHAMBER DATA UNTIL EQUILIBRIUM"
2520 PRINT " *** IS ACHIEVED "
2530 REM *****

2540 TIM%(6) = 1
2550 CALL KDTIMER'(TIM%(), "NT", "")
2560 REM CALL BGWRITE'("MBVCLOSE%", MBV$, 1, 1, "NT", "MBVSTAT3")
2570 CALL KDTIMERRD'(TM#())
2580 IF TM#(6)*IR% < 2*DTSV GOTO 2570
2590 'CALL BGWRITE'("AOVOPEN%", AOV$, 1, 1, "NT", "MBVSTAT3")
2600 CALL BGWRITE'("LMHOFF%", LMH$, 1, 1, "NT", "MBVSTAT3")
2610 CALL BGWRITE'("UWHOFF%", UWH$, 1, 1, "NT", "MBVSTAT3")
2620 CALL BGHALT'("UWHDAT4A,UWHDAT4B,UWHDAT4C,UWHDAT4D", "NT", "HALT4")
2630 CALL BGHALT'("LDATA3A,LDATA3B,LDATA3C", "NT", "HALT5")
2640 TIM%(7) = 1
2650 CALL KDTIMER'(TIM%(), "NT", "")

2660 REM
2670 PRINT " ***** SETTING UP TO ACQUIRE DATA UNTIL EQUILIBRIUM IS ATTAINED"
2680 PRINT " ***** IN THE UPPER CHAMBER"

2690 REM ***** ACQUIRE UPPER DATA *****
2700 CALL BGREAD'("UWHEAT5A", NDP5!, CJUWLWT$, BINT5%, "NONE", 1, "NT", "UWHDAT5A")
2710 CALL BGREAD'("UWHEAT5B", NDP5!, CJUGT$, BINT5%, "NONE", 1, "NT", "UWHDAT5B")
2720 CALL BGREAD'("UWHEAT5C", NDP5!, CJUFT$, BINT5%, "NONE", 1, "NT", "UWHDAT5C")
2730 CALL BGREAD'("UWHEAT5D", NDP5!, UGP$, BINT5%, "NONE", 1, "NT", "UWHDAT5D")

2740 REM ***** ACQUIRE LOWER DATA *****
2750 CALL BGREAD'("LDAT4A", NDP5!, CJLGT$, BINT5%, "NONE", 1, "NT", "LDATA4A")
2751 CALL BGREAD'("LDAT4B", NDP5!, CJLMT$, BINT5%, "NONE", 1, "NT", "LDATA4B")
2752 CALL BGREAD'("LDAT4C", NDP5!, LGP$, BINT5%, "NONE", 1, "NT", "LDATA4C")
2753 CALL BGREAD'("LDAT4D", NDP5!, CJLFT$, BINT5%, "NONE", 1, "NT", "LDATA4D")

2760 LP5! = 0!
2762 LPO! = LP5!
2764 CALL ARLASTP'("LDAT4C", LP5!)
2768 IF ABS(LPO! - LP5!) < .1 GOTO 2762
2769 CALL KDTIMERRD'(TM#())

2770 REM ***** GET UPPER DATA *****
2771 CALL ARGET'("UWHEAT5A", LP5!, LP5!, UWT$, 1, UWH1A!(), "C.THCU.K")
2772 CALL ARGET'("UWHEAT5A", LP5!, LP5!, LWT$, 1, UWH1B!(), "C.THCU.E")
2773 CALL ARGET'("UWHEAT5B", LP5!, LP5!, UGT$, 1, UWH1C!(), "C.THCU.E")
2774 CALL ARGET'("UWHEAT5C", LP5!, LP5!, UFT$, 1, UWH1D!(), "C.THCU.K")
2775 CALL ARGET'("UWHEAT5D", LP5!, LP5!, UGP$, 1, UWH1E!(), "C.VOLTS")

```

```

2776 REM ***** GET LOWER DATA *****
2777 CALL ARGET'("LDAT4A",LP5!,LP5!,LGT$,1,LGH1A!(), "C.THCU.K")
2778 CALL ARGET'("LDAT4B",LP5!,LP5!,LMT$,1,LGH1B!(), "C.THCU.K")
2779 CALL ARGET'("LDAT4C",LP5!,LP5!,LGP$,1,LGH1C!(), "C.VOLTS")
2780 CALL ARGET'("LDAT4D",LP3!,LP3!,LFT$,1,LGH1D!(), "C.THCU.K")
2781 TIME5=TM#(7)*IR%/1000

2782 PRINT "UWT,LWT,UGT,UFT,LMT,UGP ARE ", UWH1A(0), UWH1B(0), UWH1C(0),
UWH1D(0),LGH1B!(0), UWH1E(0)
2783 UFT1=UWH1D!(0): LWT1=UWH1B!(0): UWT1=UWH1A!(0):UWAVG=0.5*(UWT1+LWT1)
2784 UGP1=UWH1E!(0)*20*6.8948E-2: LGP1=LGH1C!(0)*20*6.8948E-2: UGT1=UWH1C!(0)
2785 LGT1=LGH1A!(0): LMT1=LGH1B!(0): LFT1=LGH1D!(0):UFT1=UWH1D!(0)
2786 PRINT "UWT,LWT,UGT,UFT,UGP ARE ";UWT1;LWT1;UGT1;UFT1;UGP1;"bar"
2787 PRINT "LFT1;LMT1;LGT1;LGP1";LFT1;LMT1;LGT1;LGP1;" bar"

2788 IF( UFT1>UFSETPT+UFDT AND UHST%=1) THEN GOSUB 4000
2789 IF( UFT1<UFSETPT-UFDT AND UHST%=0) THEN GOSUB 5000
2790 IF( LFT1>LFSETPT+LFDT AND LHST%=1) THEN GOSUB 4500
2791 IF( LFT1<LFSETPT-LFDT AND LHST%=0) THEN GOSUB 5500

2792 EQUIL(LP5,1)=TIME5: EQUIL(LP5,2)=UWH1C!(0) :EQUIL(LP5,3)=LGH1A!(0):
EQUIL(LP5,4)=UWH1A!(0)
2793 EQUIL(LP5,5)=UWH1B!(0): EQUIL(LP5,6)=LGH1B!(0): EQUIL(LP5,7)=UWH1E!(0):
EQUIL(LP5,8)=LGH1C!(0)

2794 REM PRINT DATA *****
2795 PRINT "TIME5,UWT,LWT,UGT,UFT,LGT,LMT,UGP,LGP
ARE",TIME5,UWH1A(0),UWH1B(0),UWH1C(0),UWH1D(0),LGH1A!(0),LGH1B!(0),UWH1E(0),LGH1C!(0)
2796 REM

2797 CALL KDTIMERRD'(TM#())
2798 IF TM#(7)*IR%/1000 < EQT GOTO 2762

2800 REM
2802 PRINT " REACTION IS OVER FOLKS! "
2810 PRINT " *****"
2820 PRINT " * TRANSFERRIN ALL KDAC DATA ARRAYS TO DISK *"
2830 PRINT " *****"
2840 REM
2841 REM ***** REACTION IS OVER

2842 SR1% = BINT1% * IR% / 1000
2843 CALL ARSAVE'("UWHEAT1A","UWT1.DAT","C.THCU.K","FT.ASCII",SR1%,"SEC")
2844 CALL ARSAVE'("UWHEAT1B","UGT1.DAT","C.THCU.E","FT.ASCII",SR1%,"SEC")
2845 CALL ARSAVE'("UWHEAT1C","UFT1.DAT","C.THCU.K","FT.ASCII",SR1%,"SEC")
2846 CALL ARSAVE'("UWHEAT1D","UGP1.DAT","C.VOLTS","FT.ASCII",SR1%,"SEC")
2847 CALL ARSAVE'("LDAT0A","LGT1.DAT","C.THCU.K","FT.ASCII",SR1%,"SEC")
2848 CALL ARSAVE'("LDAT0B","LMT1.DAT","C.THCU.K","FT.ASCII",SR1%,"SEC")
2849 CALL ARSAVE'("LDAT0C","LGP1.DAT","C.VOLTS","FT.ASCII",SR1%,"SEC")
2850 CALL ARSAVE'("LDAT0D","LFT1.DAT","C.THCU.K","FT.ASCII",SR1%,"SEC")

2900 REM

```

```

2902 SR2% = BINT2% * IR% / 1000
2910 CALL ARSAVE('UWHEAT2A',"UWT2.DAT","C.THCU.K","FT.ASCII",SR2%,"SEC")
2920 CALL ARSAVE('UWHEAT2B',"UGT2.DAT","C.THCU.E","FT.ASCII",SR2%,"SEC")
2930 CALL ARSAVE('UWHEAT2C',"UFT2.DAT","C.THCU.K","FT.ASCII",SR2%,"SEC")
2940 CALL ARSAVE('UWHEAT2D',"UGP2.DAT","C.VOLTS","FT.ASCII",SR2%,"SEC")
2950 CALL ARSAVE('LDAT1A',"LGT2.DAT","C.THCU.K","FT.ASCII",SR2%,"SEC")
2960 CALL ARSAVE('LDAT1B',"LMT2.DAT","C.THCU.K","FT.ASCII",SR2%,"SEC")
2970 CALL ARSAVE('LDAT1C',"LGP2.DAT","C.VOLTS","FT.ASCII",SR2%,"SEC")
2972 CALL ARSAVE('LDAT1D',"LFT2.DAT","C.THCU.K","FT.ASCII",SR2%,"SEC")

2980 REM
2982 SR3% = BINT3% * IR% / 1000
2990 CALL ARSAVE('UWHEAT3A',"UWT3.DAT","C.THCU.K","FT.ASCII",SR3%,"SEC")
3000 CALL ARSAVE('UWHEAT3B',"UGT3.DAT","C.THCU.E","FT.ASCII",SR3%,"SEC")
3010 CALL ARSAVE('UWHEAT3C',"UFT3.DAT","C.THCU.K","FT.ASCII",SR3%,"SEC")
3020 CALL ARSAVE('UWHEAT3D',"UGP3.DAT","C.VOLTS","FT.ASCII",SR3%,"SEC")
3030 CALL ARSAVE('LDAT2A',"LGT3.DAT","C.THCU.K","FT.ASCII",SR3%,"SEC")
3040 CALL ARSAVE('LDAT2B',"LMT3.DAT","C.THCU.K","FT.ASCII",SR3%,"SEC")
3050 CALL ARSAVE('LDAT2C',"LGP3.DAT","C.VOLTS","FT.ASCII",SR3%,"SEC")
3052 CALL ARSAVE('LDAT2D',"LFT3.DAT","C.THCU.K","FT.ASCII",SR3%,"SEC")

3060 REM
3062 SR4% = BINT4% * IR% / 1000
3070 CALL ARSAVE('UWHEAT4A',"UWT4.DAT","C.THCU.K","FT.ASCII",SR4%,"SEC")
3080 CALL ARSAVE('UWHEAT4B',"UGT4.DAT","C.THCU.E","FT.ASCII",SR4%,"SEC")
3090 CALL ARSAVE('UWHEAT4C',"UFT4.DAT","C.THCU.K","FT.ASCII",SR4%,"SEC")
3100 CALL ARSAVE('UWHEAT4D',"UGP4.DAT","C.VOLTS","FT.ASCII",SR4%,"SEC")
3110 CALL ARSAVE('LDAT3A',"LGT4.DAT","C.THCU.K","FT.ASCII",SR4%,"SEC")
3120 CALL ARSAVE('LDAT3B',"LMT4.DAT","C.THCU.K","FT.ASCII",SR4%,"SEC")
3130 CALL ARSAVE('LDAT3C',"LGP4.DAT","C.VOLTS","FT.ASCII",SR4%,"SEC")
3132 CALL ARSAVE('LDAT3D',"LFT4.DAT","C.THCU.J","FT.ASCII",SR4%,"SEC")

3140 REM
3142 SR5% = BINT5% * IR% / 1000
3150 CALL ARSAVE('UWHEAT5A',"UWT5.DAT","C.THCU.K","FT.ASCII",SR5%,"SEC")
3160 CALL ARSAVE('UWHEAT5B',"UGT5.DAT","C.THCU.E","FT.ASCII",SR5%,"SEC")
3170 CALL ARSAVE('UWHEAT5C',"UFT5.DAT","C.THCU.K","FT.ASCII",SR5%,"SEC")
3180 CALL ARSAVE('UWHEAT5D',"UGP5.DAT","C.VOLTS","FT.ASCII",SR5%,"SEC")
3190 CALL ARSAVE('LDAT4A',"LGT5.DAT","C.THCU.K","FT.ASCII",SR5%,"SEC")
3200 CALL ARSAVE('LDAT4B',"LMT5.DAT","C.THCU.K","FT.ASCII",SR5%,"SEC")
3210 CALL ARSAVE('LDAT4C',"LGP5.DAT","C.VOLTS","FT.ASCII",SR5%,"SEC")
3212 CALL ARSAVE('LDAT4D',"LFT5.DAT","C.THCU.K","FT.ASCII",SR5%,"SEC")

3220 PRINT
3230 PRINT " LITHIUM-LEAD TEST NO ", ITESTNO, " IS OVER"
3240 PRINT " RELOAD THE LOWER CHAMBER AND BEGIN A NEW TEST"
3250 PRINT
3260 PRINT "TRANSFERRING DATA ARRAYS TFOR PHASES 3,4 AND 5 TO DISK FILE"
3270 REM

3280 OPEN "COUNTD.DAT" FOR OUTPUT AS #1
3290 OPEN "REACTION.DAT" FOR OUTPUT AS #2

```


3300 OPEN "EQUIL.DAT" FOR OUTPUT AS #3

3310 FOR I= 1 TO LP3

3320 PRINT #1, CDOWN(I,1); CDOWN(I,2); CDOWN(I,3); CDOWN(I,4);CDOWN(I,5); CDOWN(I,6);
CDOWN(I,7);CDOWN(I,8)

3330 NEXT I

3340 FOR I= 1 TO LP4

3 3 5 0 P R I N T # 2 ,
REACTT(I,1);REACTT(I,2);REACTT(I,3);REACTT(I,4);REACTT(I,5);REACTT(I,6);REACTT(I,7);REACT
T(I,8)

3360 NEXT I

3370 FOR I= 1 TO LP5

3 3 8 0 P R I N T # 3 ,
EQUIL(I,1);EQUIL(I,2);EQUIL(I,3);EQUIL(I,4);EQUIL(I,5);EQUIL(I,6);EQUIL(I,7);EQUIL(I,8)

3390 NEXT I

3460 PRINT " THAT'S THE END OF DATA ACQUISITION, FOLKS!"

3470 CALL INTOFF

3480 END

4000 CALL BGWRITE'("UWHOFFF%",UWH\$,1,1,"NT","ST1")

4100 UHST%=0

4200 RETURN

4500 CALL BGWRITE'("LMHOFFF%",LMH\$,1,1,"NT","ST2")

4600 LHST%=0

4700 RETURN

5000 CALL BGWRITE'("UWHONN%",UWH\$,1,1,"NT","ST1")

5100 UHST%=1

5200 RETURN

5500 CALL BGWRITE'("LMHONN%",LMH\$,1,1,"NT","ST1")

5600 LHST%=1

5700 RETURN

END

**DATE
FILMED**

9/3/93