DECLASSIFIED

AEC-3344

HW-22765

Chemistry - General

THE COULOMETRIC DETERMINATION OF URANIUM

By

W. N. Carson, Jr.
Analytical Research
Analytical Unit

December 3, 1951

CLASSIFICATION CANCELLED

DATE 3-24-52
For The Atomic Energy Commission
William A. Stinson
Chief, Declassification Branch

NUCLEONICS DIVISION
RICHLAND, WASHINGTON

Operated for the Atomic Energy Commission by the
General Electric Company under Contract # W-31-109-Eng-52

DECLASSIFIED

U-114-1

THE COULOMETRIC DETERMINATION OF URANIUM

INTRODUCTION

The determination of uranium by titration is usually accomplished by the oxidation or reduction to the six or four valence state of the element. Of the several redox methods, one of the more useful is the oxidation of uranium(IV) to uranium(VI) by ferric ion. This method has been described by Shipley (8). It consists of pretreating the sample to remove or nullify interferences, reducing the uranium to a mixture of the U(III) and U(IV) valence states by means of a Jones reductor, and titrating with standard ferric sulfate solution back to U(VI) at an elevated temperature.

A study was made to adapt this method to a microcoulometric method. Many modifications were found necessary, but the basic titration scheme of preliminary reduction of the uranium followed by oxidation with ferric ion proved feasible and was retained. The method proved readily adaptable for use on an automatic titrator reported previously (1).

SUMMARY

An automatic coulometric titration has been developed. The method uses the automatic titrator described previously and depends upon the use of a lead reductor to reduce the uranium to the (IV) valence state, and electrolytically generated bromine in the presence of iron to oxidize the uranium to the (VI) valence state. Samples of 0.01 to

7 mg. uranium content can be automatically titrated. The interferences with the method are few, and dilutions of 10^{-5} M uranium can be titrated.

EQUIPMENT

The constructional details of the titration cell used in manual and automatic titrations are given in Figure 1. The indicator system is that of Reilley, Cooke, and Furman (6). The electrodes are two platinum

~~SECRET~~

-5-

HW-22765

wires dipping in the sample. These are polarized by a small constant current of 1.1 to 1.3 microamperes, as explained by Reilley, et. al. (6). The constant current source is composed of several 45 volt "B" batteries (Burgess Z3ONX) connected in parallel with 40 megohms resistance in series.

The generator system is a two compartment cell composed of a platinum anode immersed in the sample solution contained in one compartment, and a platinum cathode immersed in 1:1 sulfuric acid in the other compartment, the two compartments connected by means of a silica gel salt bridge. The salt bridge is prepared by filling a tube to the required length with sodium silicate solution (40° Be). The tube is immersed in 1:1 sulfuric acid, and 1:1 sulfuric acid is placed on top of the solution. The acid causes a thin film of gel to form immediately which seals in the solution. The bridge can be used after a short (15 min.) aging period. Diffusion of acid from the catholyte causes all of the solution to gel in a few days to a compact, low resistance plug. If made properly, the bridge permits only a very small flow of catholyte, and remains operative under the heat and acid conditions of the titration with a life under continuous use of several months. The catholyte must be replenished from time to time with 1:1 sulfuric acid.

The cell is heated by a small heating element (Hotspotter Heating Element, Fisher Scientific Co., Cat. No. 11-502-15) controlled by a simplytrol pyrometer (Cat. No. 4533, Assembly Products Co., Chagrin Falls, Ohio), which uses the iron-constantan thermocouple for temperature sensing. A 100-ohm 100 watt resistance is placed in series with the heater circuit to permit slower heating and extend the heater life. The temperature of titration is $95 \pm 1^{\circ}\text{C}$ and is critical. Above this temperature the solution boils and ruins the determination. Below 92° the titration reactions are too slow for automatic titrator use, and below 90° are too slow for manual use. The condenser of the cell returns water evaporated by the sweep gas from the sample.

~~SECRET~~

U-114-3

If it is not used, the evaporative losses are very large, and the sample cannot be titrated for more than a few minutes. Rapid stirring of the sample is necessary but is not critical. The gas blanket used to prevent air oxidation is carbon dioxide, but nitrogen or other inert gas could be used. The carbon dioxide is freed of oxygen by use of the gas scrubbing train shown in Figure 2. The gas is passed through water, chromous sulfate, and water again. The first water scrubber is used to saturate the gas to lessen the evaporative loss in the chromous sulfate scrubbers, the second is to remove chromous sulfate spray. The first chromous sulfate scrubber is a modification of a gas scrubber described by Shaw(7). The action is as follows: the incoming gas passes through the jet and pulls or traps a small amount of solution into the center tube. This is forced up the top of the scrubber by the gas, and splashes back on the glass beads in the top of the scrubber. The large surface exposed to the gas gives very efficient scrubbing. The liquid drains back to the reservoir at the bottom of the scrubber, where it passes through amalgamated zinc to reduce any oxidized chromium. The solution used is 15 g. /l. chromic sulfate in 1:4 sulfuric acid and is replenished by additions from the reservoir. The second scrubber is used as an indicator of failure of operation of the first, since small amounts of oxygen cause a marked change from blue to green in the color of the solution. A flow rate of approximately 100 cc. /min. per titration cell is necessary.

The manual titration apparatus is shown as a block diagram in Figure 3. The constant current source can be any suitable source; the one used has been described by Carson and Ko (2). The timer is a Standard Electric Time Company Model SM 60 electric stopwatch, used as described previously. The potentiometer is a Leeds and Northrup No. 7655 portable potentiometer; any good student type potentiometer can be used. The indicator system requires a high impedance volt meter; a Beckman Model G or Model H-2 pH meter is satisfactory. The automatic titration equipment has been described in a previous report (1).

TITRATION CHEMISTRY

Reactions. The following equations give the reactions occurring during the analysis:

Reduction of sample in lead column:

Electrolysis:

Oxidation of sample in titration cell at 95° C

Reactions 1 and 2 require the presence of a minimum of 3 M chloride or bromide. The reduction is rapid and quantitative in either case. Chloride is unsuitable for the titrations since poorly defined endpoints are obtained. The presence of chloride in the titration does not interfere, however, when a bromide media is used. The primary anode reaction product is not accurately known, although a reasonable process is the evolution of free bromine as shown. It is possible that the iron is directly oxidized, but not very probable. Iron must be present for the oxidation to be stoichiometric with respect to uranium. The controlling reaction rate is the relatively slow reaction between ferric ion and the uranium(IV). This rate is too slow for practical titrations unless the temperature is kept above 90°C.

Interferences. A number of interferences of the titration exist, but many can be easily removed by simple treatment. Molecular

~~SECRET~~

U-114-5

~~SECRET~~

-8-

HW-22765

oxygen rapidly oxidizes the hot reduced solution and must be excluded by blanketing the titration vessel with an inert gas. Dissolved oxygen in the reagent solutions causes significant errors in the titration of 10-500 microgram quantities of uranium, although this error is negligible for larger amounts of uranium. For low levels the solutions are boiled to free them of dissolved oxygen prior to use. A resume of the study made on interferences is given in Table I. The tests were made as follows: for each element, a large excess of test reagent (50-100 times the molar equivalent of uranium) was added as a solution of the compound listed in the table to 2 mg. samples of uranium. Concentrated hydrobromic acid was added, and the sample solutions boiled down to expel any volatile reaction products. The solution was diluted with dilute hydrobromic acid, reduced in the usual manner, and titrated. For elements listed as "no interference", this first sample did not change the titre of the standard sample. Several checks were made on each element at the high level if it did not give an interference. For interfering elements, lesser amounts were tried to find the maximum amount that could be tolerated. Of the interfering elements, only chromium and phosphate can be tolerated in amounts greater than traces. In the case of chromium, the interference is removed by pretitration; in the case of phosphate no interference is observed up to a 30:1 mole ratio (phosphate to uranium). Above this ratio, the precision of titration decreases; above 50:1 there is a marked interference giving erratic results.

Of the elements not listed, the rare earths, alkaline earths, and alkalis will not interfere; possibilities are some of the platinum metals, tantalum, tellurium, thorium, and niobium.

Iodine interferes in a unique manner by depolarizing the indicator electrodes whether present as iodide or free iodine. Any generated free iodine should oxidize iron at the elevated temperature, similarly to free bromine, and so not interfere with the titration reaction, but both polarized and potentiometric indicator electrodes are unusable in its presence.

~~SECRET~~

V-114-6

TABLE IRESUME OF INTERFERENCE STUDIES ON TITRATION OF URANIUM

<u>Element</u>	<u>Added as</u>	<u>Remarks</u>
Aluminum	AlCl_3	Does not interfere
Antimony	Na_3SbO_4	Non-stoichiometric interference, not pretitratable
Arsenic	Na_3AsO_4	Non-stoichiometric interference, not pretitratable
Boron	HBO_3	Does not interfere
Calcium	CaCl_2	Does not interfere
Chlorine	HCl	Does not interfere
Chromium	$\text{N}_2\text{Cr}_2\text{O}_7$	Non-stoichiometric interference, pretitratable
Cobalt	CoCl_2	Does not interfere
Copper	CuSO_4	Does not interfere
Fluorine	NaF	Non-stoichiometric interference, removed by complexing with Zr or B
Gold	AuCl_3	Does not interfere
Iodine	KI	Renders indicator system used inoperative
Magnesium	MgCl_2	Does not interfere
Manganese	KMnO_4	Does not interfere
Mercury	HgCl_2	Does not interfere
Molybdenum	Na_2MoO_4	Non-stoichiometric interference, not pretitratable
Nickel	NiCl_2	Does not interfere
Nitrogen	HNO_3	Does not interfere
Palladium	PdCl_2	Does not interfere
Phosphorus	NaH_2PO_4	Does not interfere up to 30:1 mole ratio $\text{PO}_4:\text{U}$
Platium	H_2PtCl_6	Non-stoichiometric interference, not pretitratable
Silicon	Na_2SiO_3	Does not interfere
Sulfur	Na_2SO_4	Does not interfere

~~SECRET~~

V-114-7

TABLE I (Cont.)

<u>Element</u>	<u>Added as</u>	<u>Remarks</u>
Tin	SnCl_4	Non-stoichiometric interference, not pretitratable
Titanium	TiCl_3	Non-stoichiometric interference, not pretitratable
Tungsten	Na_2WO_4	Stoichiometric interference, not pretitratable
Vanadium	NaVO_3	Non-stoichiometric interference, not pretitratable
Zinc	ZnCl_2	Does not interfere
Zirconium	ZrOCl_2	Does not interfere
Formate	HCOOH	Interferes, removed by oxidation with bromine prior to reduction
Oxalate	$\text{Na}_2\text{C}_2\text{O}_4$	Interferes, removed by oxidation with bromine prior to reduction
Sulfide	Na_2S	Interferes, removed by oxidation with bromine prior to reduction
Sulfite	Na_2SO_3	Interferes, removed by oxidation with bromine prior to reduction
NH_3	NH_4Cl	Does not interfere
Sulfamate	HSO_3NH_2	Does not interfere

The reagents used in the titration often contain significant amounts of impurities. In one case it was found the vendor's hydrobromic acid contained 0.5 mg. per liter of uranium. This level of impurity is significant because of the large amount of acid (compared to the sample) used in the determination. The iron bromide solution is a source of interference occasionally; this is best prepared by dissolving pure iron in hydrobromic acid, using free bromine to hasten the dissolution.

PROCEDURE

Reagents Used in Titration

Test Lead: C. P. granules, silver and bismuth-free.

Hydrobromic Acid: concentrated, C. P. 48%.

Electrolytic Iron: Standard Sample Company, Ames, Iowa, hydrogen fired standard electrolytic iron.

1 M Iron Bromide: Dissolve 55 g. of electrolytic iron in 500 ml. hydrobromic acid, add bromine to hasten dissolution, and dilute to 1 liter after the iron is dissolved.

Iron Rinse Solution: 0.02 M iron and 3 M HBr. Add 336 ml. concentrated HBr to 20 ml. 1 M iron bromide solution and dilute to 1 liter. Boil gently for 5 minutes just before use under a blanket of oxygen-free inert gas, if less than 500 microgram samples of uranium are used.

Sample Size. Samples containing 5 micrograms to approximately 7 mg. of uranium can be titrated by the manual titration; the lower limit of the automatic procedure is approximately 10 micrograms. The precision of the manual titration is better than that of the automatic titration for samples with less than 50 micrograms, and is poorer for samples with more than 200 micrograms.

Due to the limitation on volume in the titration cell and the difficulty in rinsing the lead reductor where large sample volumes are used, the maximum sample volume is approximately 2 ml. The volume applies to the pretreated sample volume; i. e., the sample after adjustment to a 3 M HBr content. Any size sample volume can be used if it can be adjusted to this volume prior to reduction.

Pretreatment. The preparation of the sample for reduction consists of removing interferences and adjusting to 3 M HBr content. For samples containing no interferences, the treatment consists of adding concentrated hydrobromic acid until the sample is 3 M HBr. The only limitation is that the final volume must be a maximum of approx. 2ml.

for the titration cell design given here. For samples containing nitrate, dichromate, permanganate, and other easily reduced interferences, the sample is boiled with concentrated hydrobromic acid until the bromide formed is expelled. For samples containing reducing interferences, the solution is heated with hydrobromic acid and bromine. Most of the excess bromine should be removed by boiling for a few minutes prior to reduction.

The treatment for interferences that are not removed by the simple methods given above has not been studied in detail except for fluoride. Fluoride is removed as an interference by complexing with zirconium. Methods are described by Shipley (3) for removal of some of the listed interferences.

Manual Titration. Take the required sample and pretreat to remove interferences. Outgas the titration cell thoroughly with a rapid stream of inert gas. Transfer the sample quantitatively to the top of the lead reductor, and pass it through the reductor at a rate of approximately 5 ml./min. Rinse the sample vessel three times with 2-3 ml. portions of iron rinse solution, and pass rinses through the reductor, allowing the reductor to drain to the top of the lead between each rinse. Turn on the heater and stirrer after the first rinse has been added. Rinse the column once with a 3 ml. portion of iron rinse solution. The final volume in the titration cell should be 12-15 ml. Allow the sample to heat to 95°C and start the electrolysis by closing the "titrate" switch on the constant current source. Continue the electrolysis until the voltage across the indicator electrodes is 200 mv. From this point, titrate in 5 second or 0.1 minute increments. The indicator system should be allowed to reach equilibrium between increments of electrolysis. Compute the endpoint location by the interpolation formula given previously (2).

The above procedure needs a slight modification with samples of low uranium content. With samples containing less than 0.5 mg. of

~~SECRET~~

-13-

HW-22765

uranium, the iron rinse solution must be boiled for a few minutes before use to expel oxygen.

Automatic Titration Procedure. This procedure refers to the use of the automatic titrator described previously (1). Set the control of the titrator as follows:

Constant current at factor current desired
pH meter to 0-800 mv., and standardize
Trigger at 175 mv.
Time of endpoint duration at 30 seconds
Pyrometer contact at 95°C.

Prepare the sample and introduce it into the titration cell in the same manner as for the manual titration. Turn the "titrate" switch to "On". The titrator will now heat the sample to the titration temperature, titrate it to the endpoint, and turn itself off. The titration is complete when the "Timed Out" pilot turns on. Remove the titrated sample and rinse the cell thoroughly. Record the time and turn the titrate switch to "Reset". The titrator is now ready for the next sample.

Computations. The numbers of milliequivalents of uranium are computed by multiplying the current in milliamperes by the time in seconds and dividing by 96,500. This divided by two gives the number of millimoles of uranium in the sample. For samples with less than 1 mg. of uranium, a blank correction is necessary. This is determined by manual titration at low currents with the same amount of electrolyte used in the analysis and represents the time necessary to generate sufficient ferric ion to give an endpoint. It is inversely proportional to the current and can be calculated for any current, once it is known for a particular current. The use of factor currents, especially for the automatic titrator, is recommended for routine use. With their use, the operator needs only to multiply the time of titration by the factor Q to obtain the result. By a proper choice of Q , this multiplication can be made very simple. In order to compute the current necessary to

~~SECRET~~

V-114-11

give the factor Q, the following formula can be used:

$$i = \frac{F \times Q \times S}{b \times M \times 10^6}$$

where i = current in milliamperes to give the factor Q

F = Faraday

S = Sample size in microliters

b = Unit time in seconds

M = Milliequivalent weight of compound

Q = Factor (g./l. per unit time)

Thus, the factor current making 1 minute (the unit time) of titration time equivalent to 100 g./l. uranium is computed as follows:

$$Q = 100$$

$$b = 60$$

$$M = \frac{U}{2000} = \frac{238.07}{2000}$$

$$i = \frac{96,500 \times 100 \times 2000 \times S}{60 \times 238.07 \times 10^6}$$
$$= 1.3515 \text{ ma.}$$

Using the current computed in this fashion, the uranium content in g./l. of a sample would be $100 \times t$, where t is given in minutes. A titration time of 1.345 min. would be equivalent to 134.5 g./l. uranium in the sample. Other compounds of uranium could be reported if the milliequivalent weight of the compound is used in place of the $\frac{U}{2000}$ value of this example.

RESULTS

Titration of Uranium Standard. A standard solution of uranyl bromide was prepared by dissolving metallic uranium in hydrobromic acid and oxidizing the uranium to the (VI) oxidation state with bromine. The solution was made up to contain 278.9 g./l. uranium. Dilutions of 1000:1, 100:1, and 10:1 were made, and microaliquots were taken to give the samples shown in Table II. Ten aliquots of each sample size were

TABLE II

AUTOMATIC TITRATION RESULTS ON URANIUM STANDARDS

<u>Standard g./l.</u>	<u>Sample Size ml.</u>	<u>Taken, mg.</u>	<u>Found Ave., mg.</u>	<u>Precision %</u>	<u>Ave. Titration Time, min.</u>	<u>Blank, min.</u>	<u>Current, ma.</u>
278.9	25	6.98	6.99	± 0.29	9.444	0	10.00
278.9	20	5.58	5.60	± 0.31	7.589	0	10.00
27.89	100	2.79	2.80	± 0.38	5.412	0	7.00
27.89	50	1.40	1.41	± 0.48	3.192	0.020	6.00
27.89	20	0.558	0.555	± 0.50	3.798	0.050	2.00
2.789	100	0.279	0.278	± 0.92	3.861	0.100	1.00
2.789	50	0.140	0.137	± 2.8	3.674	0.188	0.533
2.789	20	0.0558	0.0536	± 3.0	4.014	0.500	0.208
0.2789	100	0.0279	0.0290	± 5.6	2.402	0.500	0.208
0.2789	50	0.0140	0.0157	± 17.8	1.588	0.500	0.200

U-114-13

HW-22765

analyzed by the automatic titrator. The precision values in the table are for the individual determinations and were computed as follows:

$$\% \text{ Precision} = \pm \frac{\sigma}{\bar{x}} \times 100$$

where σ = best estimate of standard deviation

$$\sigma = \sqrt{\frac{\sum (x - \bar{x})^2}{n - 1}}$$

$\bar{x}$ = average value

The deviations of the average found values from the standard value were tested by Student's "t" test on the 99% confidence level, and are not significant.

In conjunction with some special problems, several other uranium standards were analyzed. Table III gives the results of these analyses. The precision is calculated as above. The samples were prepared as follows: Carefully purified uranium nitrate hexahydrate was calcined to give the oxide. This oxide was analyzed by macro gravimetric and volumetric methods a large number of times to give the result shown as the accepted value. The UO_2Cl_2 solutions were made by taking weighed portions of the oxide and dissolving in dilute hydrochloric acid, using bromine to oxidize the uranium to (VI) state. The solutions were made up by weight to the value shown in the table. For the analysis, two uranium oxide samples of approximately 0.25 g. each were dissolved in approximately 60 g. of 3 M HCl with the aid of bromine. Three aliquots of 0.5 g. of each solution were analyzed, and the results combined to give the value in the table. For the UO_2Cl_2 (Solution I), the two samples were diluted 10:1 by weight, and 0.5 g. samples of the dilution were taken as above. Six 0.5 g. aliquots of the final solution were used. For all three standards the uranium titrated per sample was from 1 to 1.5 mg.

TABLE IIIANALYSES OF STANDARD URANIUM SAMPLES

<u>Sample</u>	<u>Accepted Value</u>	<u>Found</u>	<u>Precision</u>	<u>No. Det'ns.</u>
U_3O_8	84.62% U	84.46%	$\pm 0.28\%$ U	6
UO_2Cl_2 (Solution I)	21.56 mg. U/g.	21.56 mg. U/g.	± 0.18 mg. U/g.	6
UO_2Cl_2 (Solution II)	2.538 mg. U/g.	2.559 mg. U/g.	± 0.26 mg. U/g.	6

The method has been used in the analysis of some special samples, and the results are reported by E. M. Kinderman in another report (4). The method has also been used in the analysis of solutions of uranium as low as 10^{-5} M. The results are reported by E. M. Kinderman in another report (5).

DISCUSSION

The differential polarographic indicator system used behaves as expected for the irreversible uranyl-uranous couple by developing a high potential at the start of the titration, which decreases sharply at the endpoint. The stoichiometric endpoint in our apparatus occurs at 175 mv. The endpoint shifts for different polarization currents and size of electrodes, but can be found definitely by titration in increments across the endpoint. The setting of the trigger cut-off potential at the endpoint is not critical, as a change in 50 mv. will not result in appreciable change in titration time.

During the development work on the method, a potentiometric indicator system, using calomel-platinum electrodes, was used. This was satisfactory insofar as behavior of the system was concerned, but

frequent failures of the calomel electrode were experienced, caused by the high temperature of operation. The endpoint potential in this system is 260 mv. This system would be usable if a reference electrode suitable for use in the hot acid solutions were developed.

Both the potentiometric and differential polarographic indicators show a marked anticipatory behavior. This is because the ferric iron generated does not react rapidly with the uranium, and a small excess concentration of ferric ion can be present during the titration. As stated previously, this behavior is an asset in making the titration automatic, as it prevents overtitration.

Ferric ion exists as a highly colored bromide complex in the titration media. The endpoint can be detected by observing the appearance of this dark color, if no other colored ions are present. This color formation is not suitable for precise work, however, for the appearance of a darkly colored solution at the end of a titration is a sign of overtitration, if the original solution is light in color.

The autotitrator used was set up in duplex, i. e., two titration cells were used for one power supply. The operator can thus prepare a sample while the previous one is being titrated. Without undue exertion by the operator, 35-40 samples can be titrated per working day. This is in contrast to the average of 12-15 per day when the titration is performed manually.

The primary difficulty in the method lies in the removal of interferences. Traces of nitrate, oxygen, and other oxidizing agents are not reduced by the lead reductor, but oxidize the iron in the hot hydrobromic acid medium, causing low results. Therefore, it is imperative that all samples which may have such traces of oxidants be pretreated by heating with hydrobromic acid before reduction.

The lead reductor is relatively unaffected by most foreign ions, except in the case where the ion tends to be reduced and plated out. This

is true of copper and tin. Phosphate, sulfate, and other anions which form insoluble lead salts cause no difficulties as long as the solutions are 3 M in hydrobromic acid. The reductor tends to retain small amounts of uranium unless it is rinsed with a dilute iron solution. No retention is observed when the iron solution is 0.02 M or above. Hot acid solutions make the reductor gas appreciably but do not affect its operation. Air does not cause peroxide formation as in the Jones reductor, so accidental admission of air to the column is not harmful.

The iron concentration affects the endpoint such that the lower the concentration the sharper the endpoint. The lower limit is set by the rinsing requirements. The potentiometric indicator system responds only to the ferric-ferrous couple. For the differential polarographic system the potential is also independent of uranium over a large range of concentration.

The method appears to be most useful in the titration of solutions of uranium which do not have impurities that are not removed by the simple pretreatment of boiling with an excess of acid or with free bromine. The wide range of concentrations that can be easily handled is also useful. Finally, the method is easily learned and relatively free of operator error.

ACKNOWLEDGMENTS

The author wishes to acknowledge the work on this method performed by some of his co-workers. These were D. J. Watkins, who did the precision studies, many of the interference studies, and helped in the differential polarographic endpoint studies; E. M. Kinderman, who helped with the pretreatment and interference studies and the early automatic titrations; W. W. Mills, who furnished many of the standards for the investigation; and R. Ko and G. Meyer, who established the need for iron in the generation system.

BIBLIOGRAPHY

- (1) Carson, W. N., Jr., "An Automatic Titrator Using Electrolytically Generated Titrants," HW-22780 (Secret).
- (2) Carson, W. N., Jr. and Ko, R., Anal. Chem. 23, 1019 (1951).
- (3) Cooke, W. D., Hazel, F., and McNabb, W. M., Anal. Chem. 22, 654 (1950).
- (4) Kinderman, E. M., "Coulometric Method of Analysis Applied to Determination of Uranium in Aluminum-Uranium Solution," HW-22782 (Secret).
- (5) Kinderman, E. M., "Determination of Uranium in Aluminum-Silicon Alloys," HW-22781 (Secret).
- (6) Reilley, C. N., Cooke, W. D., and Furman, N.H., Anal. Chem. 22, 1223 (1951).
- (7) Shaw, J. A., Ind. Engr. Chem. Anal. Ed. 6, 479 (1934).
- (8) Shipley, E. D., AECD-2804.

WNC:dt

W. N. Carson Jr.
W. N. Carson, Jr.

~~SECRET~~

U-114-18

FIGURE 1
TITRATION CELL

FIGURE 4
GAS SCRUBBING TRAIN

U-114-20

FIGURE 3
MANUAL TITRATION APPARATUS

U-114-21

END