Acoustic Probe for Solid-Gas-Liquid Suspensions

US Department of Energy Environmental Management Science Program
Grant No. DE-FG07-96ER14729

First Year Informal Report

Department of Chemical Engineering and Materials Science
Syracuse University
Syracuse, NY 13244-1190

P.I.: Professor L.L. Tavlarides

phone: 315-443-1883

fax: 315-443-2559

Co-P.I.: Professor AS. Sangani

Collaborators: Dr. M.S. Greenwood

Battelle Pacific Northwest National Laboratory

Mr. Michael A. Norato

Dr. Peter D.M. Spelt

Introduction:

Acoustic probes have shown promise to be quite effective in determining the solid content in solid-liquid suspensions. However, the presence of small amounts of gas in the waste slurries stored in tanks across the DOE complex prevents straightforward application for characterization of these slurries. The proposed research will develop an acoustic probe for monitoring particle size and volume fraction in slurries in the absence and the presence of gas bubbles.

Theoretical Analysis Accomplished:

Attenuation of sound waves depends on the size distribution of the solids and the volume fraction of solids. These can in principle be calculated from attenuation measured over a range of frequencies. However, small amounts of bubbles distort the measured attenuation. A typical result from theoretical analysis for the attenuation of solid-gas-liquid systems is given in Figure 1.

The total attenuation of a sound wave $v(\omega)$ equals the sum of contributions by a large number of 'bins' of particle sizes. This notion yields the following equation for the (hitherto) unknown number density of solid particles as a function of particle radius $N(a)$:

$$\int k(\omega, a) N(a) da = v(\omega),$$

where the kernel $k(\omega, a)$ is obtained from analysis. If $N(a)$ is given, the above equation is used to calculate the attenuation $v(\omega)$. This is referred to as solving the 'forward problem'. Solving for $N(a)$ with $v(\omega)$ given is the 'inverse problem'. A complication that one faces when trying to solve the inverse problem is that the stated problem is mathematically ill-posed, i.e., small fluctuations in $v(\omega)$ cause large fluctuations in the result for the number density. Therefore the problem needs to be 'regularized', i.e., the stated problem needs to be changed slightly such as to make it well-posed. This has been done by others for gas-liquid systems in the past.

This approach is currently being applied in the present project to solid-liquid systems. As is shown in Figure 2, it successfully recovers the number density that has been used in the forward problem to generate attenuation data. Having this solution technique giving reliable results for the inverse problems of both gas-liquid and solid-liquid systems, we shall apply this method in the near future to solid-gas-liquid systems.

Experimental Work Accomplished:

Experimentation has initially proceeded with solid-liquid systems. The experimental setup consists of a desktop computer containing a Matec TB-1000 Toneburst Card (as shown in Figure 3), a mixing vessel equipped with Panametrics videoscans transducers, an attenuator box, and a LeCroy Model 9310A digital Oscilloscope. The system is operated in pitch/catch mode. Toneburst signals of various frequencies are sent from the toneburst card to the sending transducer (simultaneously triggering the scope), propagate through the sample medium to the receive transducer. Then, the signal is sent either directly to the oscilloscope, or to the receiver/amplifier in the toneburst card and then to the oscilloscope. In either case the received signal is displayed on the oscilloscope where the amplitude and travel time are measured. Six pairs of transducers are used to span a frequency range from 0.2 to 12.0 MHz. Attenuation is calculated from the voltage measurements.

We have experimented with solid-liquid slurries of soda-lime glass beads (0.5-1 8.5 μm in radius, 9.5 μm mean) in water and obtained the results as shown in Figure 4 for slurries of 5% and 10% by volume of solids. In Figure 4 the experimental results are compared with results of the theory for the forward problem, and the agreement is quite good considering that the exact particle size distribution of the particles was unknown and had to be assumed in the theoretical calculations. The theoretical results assume a Gaussian distribution of particle radius between 1.0 μm and 21.0 μm , with a mean radius of 10 μm and standard deviation of 5 μm .

Experiments are currently underway to examine the possibility of employing Fast Fourier Transforms (FFT) in the analysis of the received signals of spike pulse input signals which are transmitted through slurries. The use of FFT's will allow attenuation data as a function of frequency to be acquired much more rapidly than time domain attenuation data.

Attenuation measurements are also being made in bubbly liquids, and various methods of bubble generation are being examined. Once the FFT method is refined for solid-liquid systems and bubbly liquids, attenuation measurements in three phase systems will be performed.

Figure 1. Attenuation ν as a function of frequency ω in a solid-gas-liquid system; result from analysis. Solids volume fraction is 30%, bubbles volume fraction is 0.1%. Radius of both solids and bubbles is 10 microns. The peak between 0.1 and 1 MHz is caused by the presence of bubbles; the general increase of attenuation with frequency is caused by the solids.

Figure 2. Comparison between the exact number density of solids in a solid-liquid system (continuous line) and its predicted value obtained from solving the inverse problem (dashed line). Solids volume fraction is 5%.

Figure 3. Experimental Setup for attenuation measurements in solid-liquid systems.

Figure 4. Comparison of forward problem theory results and experimental solid-liquid attenuation measurements at 5% and 10% solids by volume. The experimental solids radii range from 0.5-18.5 μm with 9.5 μm mean. The theoretical results are for a Gaussian distribution of particles with solids radii ranging from 1.0 μm and 21.0 μm , with a mean radius of 10 μm and standard deviation of 5 μm .