ENGINEERING CHANGE NOTICE

Page 1 of 2

1. ECN 633339

Proj.
ECN

2. ECN Category (mark one) Supplemental ☐ Direct Revision ☒ Change ECN ☐ Temporary ☐ Standby ☐ Supersedure ☐ Cancel/Void ☐	3. Originator's Name, Organization, MSIN, and Telephone No. Franciska H. Steen, Production Planning & Control, T6-06, 372-2495	4. USQ Required? ☐ Yes ☒ No	5. Date 01/21/97
12a. Modification Work ☐ Yes (fill out Blk. 12b) ☒ No (NA Blks. 12b, 12c, 12d)	12b. Work Package No. N/A	12c. Modification Work Complete N/A Design Authority/Cog. Engineer Signature & Date	12d. Restored to Original Condition (Temp. or Standby ECN only) N/A Design Authority/Cog. Engineer Signature & Date
13a. Description of Change This ECN is being generated in order to update/add additional information to the existing document.			
14a. Justification (mark one) Criteria Change ☒ Design Improvement ☐ Environmental ☐ Facility Deactivation ☐ As-Found ☐ Facilitate Const ☐ Const. Error/Omission ☐ Design Error/Omission ☐			
14b. Justification Details This ECN will assist in the completion of the documentation process for this document.			
15. Distribution (include name, MSIN, and no. of copies) See attached distribution.			

RELEASE STAMP

DATE

GTA 4

JAN 24 1997

ENGINEERING CHANGE NOTICE

Page 2 of 2

1. ECN (use no. from pg. 1)

ECN-633339

16. Design Verification Required ☐ Yes ☒ No	17. Cost Impact ENGINEERING Additional ☐ \$ Savings ☐ \$	CONSTRUCTION Additional ☐ \$ Savings ☐ \$	18. Schedule Impact (days) Improvement ☐ Delay ☐
19. Change Impact Review: Indicate the related documents (other than the engineering documents identified on Side 1) that will be affected by the change described in Block 13. Enter the affected document number in Block 20.			
SDD/DD	☐	Seismic/Stress Analysis	☐
Functional Design Criteria	☐	Stress/Design Report	☐
Operating Specification	☐	Interface Control Drawing	☐
Criticality Specification	☐	Calibration Procedure	☐
Conceptual Design Report	☐	Installation Procedure	☐
Equipment Spec.	☐	Maintenance Procedure	☐
Const. Spec.	☐	Engineering Procedure	☐
Procurement Spec.	☐	Operating Instruction	☐
Vendor Information	☐	Operating Procedure	☐
OM Manual	☐	Operational Safety Requirement	☐
FSAR/SAR	☐	IEFD Drawing	☐
Safety Equipment List	☐	Cell Arrangement Drawing	☐
Radiation Work Permit	☐	Essential Material Specification	☐
Environmental Impact Statement	☐	Fac. Proc. Samp. Schedule	☐
Environmental Report	☐	Inspection Plan	☐
Environmental Permit	☐	Inventory Adjustment Request	☐
20. Other Affected Documents: (NOTE: Documents listed below will not be revised by this ECN.) Signatures below indicate that the signing organization has been notified of other affected documents listed below. <div style="display: flex; justify-content: space-between;"> <div>Document Number/Revision</div> <div>Document Number/Revision</div> <div>Document Number/Revision</div> </div>			
N/A			
21. Approvals			
Design Authority	Signature	Date	Signature Date
Cog. Eng. F. H. Steen		1/23/97	PE
Cog. Mgr. A. D. Rice		1-23-97	QA
QA T. L. Tung		1-23-97	Safety
Safety			Design
Environ.			Environ.
Other			Other
DEPARTMENT OF ENERGY			
Signature or a Control Number that tracks the Approval Signature			
ADDITIONAL			

Tank 241-AN-105, Cores 152 and 153 Analytical Results for the Final Report

Franciska H. Steen

Rust Federal Services of Hanford, Inc., Richland, WA 99352
U.S. Department of Energy Contract DE-AC06-87RL10930

EDT/ECN: ECN-633339 UC: 2070
Org Code: 7576B Charge Code: MD378
B&R Code: EW 3120074 Total Pages: 2496

Key Words: Tank 241-AN-105, Tank AN-105, AN-105, AN Farm, Core 152, Core 153, Analytical Results, Final Report

Abstract: N/A

TRADEMARK DISCLAIMER. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

Printed in the United States of America. To obtain copies of this document, contact: WHC/BCS Document Control Services, P.O. Box 1970, Mailstop H6-08, Richland WA 99352, Phone (509) 372-2420; Fax (509) 376-4989.

Release Approval

1/24/97
Date

Approved for Public Release

HNF
~~WAC~~ SD-WM-DP-199, REV. 1
KMB Y24M7

ANALYTICAL SERVICES

**TANK 241-AN-105, CORES 152 AND 153
ANALYTICAL RESULTS FOR THE FINAL REPORT**

Project Coordinator: FRANCISKA H. STEEN

**Prepared for the U.S. Department of Energy
Office of Environmental Restoration
and Waste Management**

by

**222-S Laboratory
Rust Federal Services of Hanford Inc.
P.O. Box 700
Richland, Washington**

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

TABLE OF CONTENTS

Narrative	1
AN-105 Sample Breakdown (Attachment 1)	13
Sample Data Summary	56
Photographs	164
Sample Handling	203
Extrusion Worklist # 9966 (3481)	216
Extrusion Worklist # 9967 (3482)	217
Extrusion Worklist # 9968 (3483)	218
Extrusion Worklist # 9969 (3491)	219
Extrusion Worklist # 9970 (3492)	220
Extrusion Worklist # 9971 (3493)	221
Extrusion Worklist # 9972 (3494)	222
Extrusion Worklist # 9973 (3495)	223
Extrusion Worklist # 10012 (3502)	224
Extrusion Worklist # 10014 (3503)	225
Extrusion Worklist # 10015 (3504)	226
Extrusion Worklist # 10110 (3505)	227
Extrusion Worklist # 10181 (3512)	228
Extrusion Worklist # 10182 (3513)	229
Extrusion Worklist # 10183 (3514)	230
Extrusion Worklist # 10388 (3588)	231
Extrusion Worklist # 10389 (3589)	232
Extrusion Worklist # 10390 (3590)	233
Extrusion Worklist # 10611 (3630)	234
Extrusion Worklist # 10613 (3637)	235
Extrusion Worklist # 10614 (3638)	236
Extrusion Worklist # 10615 (3639)	237
Extrusion Worklist # 10652 (3663)	238
Extrusion Worklist # 10653 (3664)	239
Extrusion Worklist # 10654 (3665)	240
Extrusion Worklist # 10655 (3666)	241
Extrusion Worklist # 11399 (4154)	242
Extrusion Worklist # 11400 (4155)	243
Extrusion Worklist # 11401 (4156)	244
Extrusion Worklist # 11693 (4324)	245
Extrusion Worklist # 11702 (4323)	246
Extrusion Worklist # 11780 (4326)	247
Extrusion Worklist # 11781 (4327)	248
Extrusion Worklist # 11782 (4328)	249
Extrusion Worklist # 11932 (4384)	250

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 11934 (4385)	251
Extrusion Worklist # 11936 (4386)	252
Sample Preparations	253
Fusion Digestion Worklist # 11117 (3827, 3893, 3894, 3895)	255
Fusion Digestion Worklist # 11146 (3889, 3890, 3896, 3897)	257
Fusion Digestion Worklist # 11147 (3891, 3892)	259
Fusion Digestion Worklist # 12107 (4314, 4317, 4337)	260
Fusion Digestion Worklist # 12108 (4441, 4442, 4443)	261
Fusion Digestion Worklist # 12109 (3984, 3985, 3986)	262
Fusion Digestion Worklist # 12383 (4256, 4261)	263
Water Digestion Worklist # 11150 (3907, 3915)	264
Water Digestion Worklist # 11151 (3829, 3908)	265
Water Digestion Worklist # 11152 (3909, 3910)	266
Water Digestion Worklist # 11153 (3911, 3912)	267
Water Digestion Worklist # 11154 (3913, 3914)	268
Water Digestion Worklist # 12025 (3990, 3991)	269
Water Digestion Worklist # 12026 (3992)	270
Water Digestion Worklist # 12113 (4316, 4339)	271
Water Digestion Worklist # 12114 (4319, 4447)	272
Water Digestion Worklist # 12115 (4448, 4449)	273
Water Digestion Worklist # 12253 (593)	274
Water Digestion Worklist # 12385 (4258, 4263)	275
Water Digestion Worklist # 13659 (5325, 5326,)	276
Acid Digestion Worklist # 11148 (3906, 3898, 3828, 3899, 3900, 3901)	277
Acid Digestion Worklist # 11149 (3902, 3903, 3904, 3905)	279
Acid Digestion Worklist # 12110 (4315, 4318, 4338, 4444)	281
Acid Digestion Worklist # 12111 (3988, 4445, 4446)	283
Acid Digestion Worklist # 12112 (3987, 3989)	285
Acid Digestion Worklist # 12252 (592)	286
Acid Digestion Worklist # 12264 (4144)	287
Acid Digestion Worklist # 12384 (4257, 4262)	288
AN-105 Bulk Density Worksheets	289
Inorganic Analyses	304
Differential scanning Calorimetry (DSC)	
DSC Worklist # 12274 (4039)	306
DSC Worklist # 12325 (4254, 4259)	311
Thermogravimetric Analyses (TGA)	
TGA Worklist # 12278 (4039)	317
TGA Worklist # 12334 (4254, 4259)	324

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

TABLE OF CONTENTS (Continued)

IC Worklist # 11955 (3913, 3914)	794
IC Worklist # 12409 (4308, 4309)	806
IC Worklist # 12410 (4333, 4419)	817
IC Worklist # 12411 (4420, 4421)	828
IC Worklist # 12412 (4433, 4434)	841
IC Worklist # 12420 (4039)	852
IC Worklist # 12473 (3744)	861
IC Worklist # 12475 (3990)	872
IC Worklist # 12476 (3992)	886
IC Worklist # 12478 (4339, 4316)	896
IC Worklist # 12479 (4319, 4447)	908
IC Worklist # 12502 (4309, 4333)	920
IC Worklist # 12503 (4419, 4420)	930
IC Worklist # 12504 (4421, 4433)	940
IC Worklist # 12505 (4434)	952
IC Worklist # 12506 (3990, 3991)	960
IC Worklist # 12508 (4339, 4316)	972
IC Worklist # 12509 (4319, 4447)	983
IC Worklist # 12683 (3747, 3739)	996
IC Worklist # 12832 (4258, 4263)	1007
IC Worklist # 13020 (4258, 4263)	1020
IC Worklist # 13023 (3744)	1031
IC Worklist # 13202 (3991)	1046
IC Worklist # 13311 (3744)	1056
IC Worklist # 13601 (4420, 4421)	1068
IC Worklist # 13814 (4448, 4449)	1078
IC Worklist # 14081 (3909, 3910)	1090
IC Worklist # 14082 (4308, 4831)	1102
IC Worklist # 14520 (5325, 5326)	1112
IC Worklist # 14704 (5325, 5326)	1123
IC Worklist # 15170 (3913, 3914, 3992)	1135
IC Worklist # 15198 (3911, 3912)	1152
IC Worklist # 15476 (4258, 4263)	1162
IC Worklist # 15852 (4419)	1174
Inductively Coupled Plasma Spectroscopy Analyses (ICP)	
ICP Worklist # 10858 (3631)	1181
ICP Worklist # 11595 (3732, 3739, 3740, 3741, 3742)	1188
ICP Worklist # 11597 (3743, 3744, 3745, 3746)	1198
ICP Worklist # 11598 (3747, 3748, 3749, 3750)	1207
ICP Worklist # 11601 (3751, 3974, 3975, 3976)	1216
ICP Worklist # 11602 (3977, 3978, 3979, 3980)	1225
ICP Worklist # 11976 (3828, 3898, 3899, 3900, 3901, 3906)	1234
ICP Worklist # 12085 (3902, 3903, 3904, 3905)	1243
ICP Worklist # 12098 (3827, 3893, 3894, 3895)	1252
ICP Worklist # 12131 (3889, 3890, 3896, 3897)	1261

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

TABLE OF CONTENTS (Continued)

ICP Worklist # 12132 (3891, 3892)	1270
ICP Worklist # 12346 (4308, 4309)	1279
ICP Worklist # 12347 (4333, 4419, 4420)	1285
ICP Worklist # 12348 (4421, 4433, 4434)	1292
ICP Worklist # 12443 (3987, 3989)	1299
ICP Worklist # 12444 (3988, 4445, 4446)	1306
ICP Worklist # 12445 (4315, 4318, 4338, 4444)	1313
ICP Worklist # 12446 (3984, 3985, 3986)	1322
ICP Worklist # 12447 (4314, 4317, 4337)	1329
ICP Worklist # 12449 (4441, 4442, 4443)	1336
ICP Worklist # 12670 (4831)	1343
ICP Worklist # 12849 (4257, 4262)	1349
Total Inorganic Carbon/Total Organic Carbon Analyses (TICTOC)	
TICTOC Worklist # 12388 (4049)	1356
TICTOC Worklist # 12402 (4254, 4259)	1376
Uranium Analyses (U)	
U Worklist # 12764 (4256, 4261)	1406
U Worklist # 14454 (4003)	1419
U Worklist # 14576 (5546)	1428
Radiochemical Analyses	
	1437
Tritium Analyses (H3)	
H3 Worklist # 12793 (4258)	1439
H3 Worklist # 14427 (4003)	1446
H3 Worklist # 14765 (4831)	1453
H3 Worklist # 15289 (4263)	1460
Total Alpha Analyses (AT)	
AT Worklist # 11366 (3743, 3744, 3745, 3746, 3747)	1466
AT Worklist # 11612 (3975, 3976, 3977, 3978, 3979, 3980)	1482
AT Worklist # 11862 (3893, 3827)	1500
AT Worklist # 11866 (3889, 3890)	1509
AT Worklist # 11977 (3732, 3739, 3740, 3741, 3742)	1518
AT Worklist # 12123 (3891, 3892)	1534
AT Worklist # 12275 (4039, 4419, 4420, 4433, 4434)	1543
AT Worklist # 12280 (4308, 4309, 4333)	1563
AT Worklist # 12461 (3985)	1575
AT Worklist # 12463 (4317, 4337)	1585
AT Worklist # 12465 (4442, 4443)	1595
AT Worklist # 12602 (3984)	1604
AT Worklist # 12603 (4421)	1611
AT Worklist # 12953 (3748, 3749, 3750, 3751)	1618

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

TABLE OF CONTENTS (Continued)

Total Alpha/Total Beta Analyses (AT/TB)	
AT/TB Worklist # 11611 (3974)	1631
AT/TB Worklist # 12752 (4256, 4261)	1643
Total Beta Analyses (TB)	
TB Worklist # 12309 (4049)	1661
Gamma Energy Analyses (GEA)	
GEA Worklist # 12312 (4049)	1668
GEA Worklist # 12759 (4256, 4261)	1687
GEA Worklist # 14451 (4003)	1714
Iodine-129 Analyses (I-129)	
I-129 Worklist # 12313 (4049)	1733
I-129 Worklist # 12795 (4258, 4263)	1752
I-129 Worklist # 14553 (4003)	1778
Strontium Analyses (Sr-90)	
Sr-90 Worklist # 12516 (4049)	1796
Sr-90 Worklist # 13052 (4256, 4261)	1804
Sr-90 Worklist # 15407 (4003)	1814
Technetium-99 Analyses (Tc-99)	
Tc-99 Worklist # 12763 (4256, 4261)	1820
Tc-99 Worklist # 13001 (4049, 414)	1830
Tc-99 Worklist # 14447 (4003)	1839
Neptunium-237 Analyses (Np-237)	
Np-237 Worklist # 12762 (4256, 4261)	1847
Np-237 Worklist # 12315 (4049)	1857
Np-237 Worklist # 14429 (4003)	1864
Americium-241 Analyses (Am-241)	
Am-241 Worklist # 12311 (4049)	1871
Am-241 Worklist # 13071 (4256, 4261)	1890
Am-241 Worklist # 14426 (4003)	1918
Plutonium Analyses (Pu)	
Pu Worklist # 12316 (4049)	1937
Pu Worklist # 12761 (4256, 4261)	1957
Pu Worklist # 14431 (4003)	1985
Part II	Tank 241-AN-105, Cores 152 and 153, Analytical Results for the 45 Day Report
Narrative	2-1

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

TABLE OF CONTENTS (Continued)

AN-105 Sample Breakdown (Attachment 1)	2-12
Sample Data Summary	2-51
Chain of Custody Forms	2-90
Inorganic Analyses	2-138
Differential Scanning Calorimetry (DSC)	
DSC Worklist # 11288 (3732, 3739)	2-141
DSC Worklist # 11289 (3740, 3741)	2-147
DSC Worklist # 11290 (3742, 3743)	2-153
DSC Worklist # 11291 (3744, 3745)	2-159
DSC Worklist # 11292 (3746, 3747)	2-165
DSC Worklist # 11293 (3748, 3749)	2-171
DSC Worklist # 11294 (3750, 3751)	2-176
DSC Worklist # 11308 (3731)	2-182
DSC Worklist # 11499 (3792, 3793)	2-186
DSC Worklist # 11500 (3794, 3795)	2-192
DSC Worklist # 11501 (3796, 3797)	2-198
DSC Worklist # 11502 (3798, 3816)	2-204
DSC Worklist # 11503 (3817)	2-210
DSC Worklist # 11658 (3974, 3975)	2-214
DSC Worklist # 11659 (3976, 3977)	2-220
DSC Worklist # 11660 (3978, 3979, 3980)	2-226
DSC Worklist # 11663 (3981, 3982, 3983)	2-235
DSC Worklist # 12274 (4039, 4308)	2-243
DSC Worklist # 12327 (4310, 4311)	2-249
DSC Worklist # 12328 (4335, 4435)	2-255
DSC Worklist # 12329 (4436, 4437)	2-261
DSC Worklist # 12330 (4309, 4333)	2-267
DSC Worklist # 12331 (4419, 4420)	2-274
DSC Worklist # 12332 (4421, 4433)	2-280
DSC Worklist # 12333 (4434)	2-287
DSC Worklist # 12356 (3731, 4254, 3792, 3793, 3794, 3795, 3796, 3797)	2-291
DSC Worklist # 12357 (3798, 3816, 3817, 4259, 3981, 3982, 3983, 4310, 4311)	2-294
DSC Worklist # 12358 (4335, 4435, 4436, 4437, 4433, 4434)	2-297
DSC Worklist # 12359 (3732, 4039, 3739, 3740, 3741, 3742, 3743, 3744, 3745)	2-299
DSC Worklist # 12360 (3746, 3747, 3748, 3749, 3750, 3751, 3974, 3975, 3976, 3977)	2-302
DSC Worklist # 12361 (3978, 3979, 3980, 4308, 4309, 4333, 4419, 4420, 4421)	2-305

TABLE OF CONTENTS (Continued)

DSC Worklist # 13019 (4419, 3749)	2-308
Thermogravimetric Analysis (TGA)	
TGA Worklist # 11281 (3732, 3739)	2-316
TGA Worklist # 11282 (3740, 3741)	2-322
TGA Worklist # 11283 (3742, 3743)	2-329
TGA Worklist # 11284 (3744, 3745)	2-335
TGA Worklist # 11285 (3746, 3747)	2-341
TGA Worklist # 11286 (3748, 3749)	2-348
TGA Worklist # 11287 (3750, 3751)	2-354
TGA Worklist # 11305 (3731)	2-361
TGA Worklist # 11493 (3792, 3793)	2-365
TGA Worklist # 11494 (3794, 3795)	2-371
TGA Worklist # 11495 (3796, 3797)	2-377
TGA Worklist # 11496 (3798, 3816)	2-383
TGA Worklist # 11497 (3817)	2-389
TGA Worklist # 11651 (3974, 3975)	2-393
TGA Worklist # 11652 (3976, 3977)	2-399
TGA Worklist # 11653 (3978, 3979, 3980)	2-405
TGA Worklist # 11656 (3981, 3982, 3983)	2-414
TGA Worklist # 12278 (4039, 4308)	2-423
TGA Worklist # 12336 (4310, 4311)	2-430
TGA Worklist # 12337 (4335, 4435)	2-436
TGA Worklist # 12338 (4436, 4437)	2-442
TGA Worklist # 12339 (4309, 4333)	2-448
TGA Worklist # 12340 (4419, 4420)	2-455
TGA Worklist # 12341 (4421, 4433)	2-461
TGA Worklist # 12342 (4434)	2-468
TGA Worklist # 12607 (3748)	2-472

This Document consists of two sections.

Part I consists of pages 1 through 2003, plus pages ii, 2, 14, 57, 165, 204, 254, 290, 305, 950, 951 and 1438 were intentionally left blank.

Part II consists of pages 2-1 through 2-475, and pages 2-2, 2-12.1, 2-52, 2-91, 2-139 and 2-140 were intentionally left blank.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc. Mettler is a Registered Trademark of Mettler Electronics.

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

NARRATIVE

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

222-S ANALYTICAL SERVICES

**TANK 241-AN-105, CORES 152 AND 153,
ANALYTICAL RESULTS FOR THE FINAL REPORT**

This document is the analytical laboratory report for tank 241-AN-105 push mode core segments collected between June 10, 1996 and June 28, 1996. The segments were subsampled and analyzed in accordance with the *Tank 241-AN-105 Push Mode Core Sampling and Analysis Plan* (TSAP) (Steen, 1997), the *Safety Screening Data Quality Objective* (DQO) (Dukelow, et al., 1995) and the *Flammable Gas Data Quality Objective* (DQO) (Benar, 1995). The analytical results are included in the data summary table (Table 1).

Sample Handling

Attachment 1 is a cross reference to relate the tank farm identification numbers to the 222-S Laboratory LABCORE/LIMS sample numbers. These diagrams show the make-up of the core composites and secondary analyses requested. These diagrams and any changes made supersede the diagrams presented in Rev. 0 of this document. Sample handling and appearance information for the subsegments may be found in Rev. 0 of this document.

Two solid and one liquid core composite were made from core 152. The core composite worksheets are included in this document.

Selected segments from core 152 (15, 17, 19 and 21) and core 153 (4, 16, 18 and 20) were sampled using the Retained Gas Sampler (RGS) and extruded by the Process Chemistry and Statistical Analysis Group. Core 152 segment 14 was originally identified as a RGS sample. However since the integrity of the sample was compromised due to a leaking sampler seal it was extruded as a non-RGS sample.

Field Blank

A field blank was provided to the 222-S laboratory with core 153. It underwent the same analysis as the drainable liquid as instructed by the TSAP (Steen, 1997).

Analytical Results Summary

The data summary report (Table 1) compiles the analytical results that satisfy all applicable DQOs.

Inorganic Analyses**Differential Scanning Calorimetry (DSC)**

The DSC analyses were performed on the core composite subsamples. The analyses were performed in duplicate on direct subsamples. The exothermic energy based on dry weight of subsample was calculated for all subsamples. The average of the TGA results for each subsample was used in the dry weight correction for that subsample. The results for all DSC analyses and the raw data for the core composite subsamples are included in this report.

High RPDs (>20%) were reported for the solid core composite subsamples. The RPDs can be attributed to the heterogenous nature of the samples and no reruns were requested. The standard recoveries for this analysis were within the required limits.

Thermogravimetric Analysis (TGA)

The TGA analyses were performed on the core composite subsamples. The analyses were performed in duplicate on direct subsamples. Results were determined by summing the weight loss steps which occurred below 220°C; weight loss steps above this were not used to determine the result. The results for all TGA analyses and the raw data for the core composite subsamples are included in this report. The standard recoveries and RPDs for this analysis were within the required limits.

Sp.G.-Specific Gravity/Bulk Density

Bulk density analyses were performed on direct solid subsamples. The results and raw data for all subsamples are reported in this revision. Specific gravity analyses were performed on the liquid subsamples. The results are provided in Table 1. The raw data are included in this revision. There are no exceptions to the quality control parameters stated in the TSAP for these samples.

Inductively Coupled Plasma Spectrophotometry (ICP)

The liquid subsamples were prepared for analysis by an acid adjustment of the direct subsample. This is indicated by a "D" in the aliquot class (A#) column in Table 1. Solid subsamples were prepared for analysis by performing both an acid digest as indicated by an "A" and a fusion as indicated by an "F".

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Only the QC for requested analytes (Li, Na, Al, Cr, Ca, Fe, K, U, Zn, Zr, Ba, Si, B, Bi, Mn, Ni) were reviewed for this report. Other "opportunistic" analyte results are included in Table 1. These analytes do not have customer defined QC parameters and are not discussed. High RPDs (>20%) were reported for several analytes (Li, Al, Cr, Ca, Fe, Zn, Zr, Si, B, Mn, Ni) and were the result of sample inhomogeneity. Reruns were not requested. Table 2 shows a comparison of the result of the sample with that of the serial dilution for those required analytes that had a spike recovery outside of the 75% to 125% range as set by the TSAP and the Quality Assurance Plan for the 222-S Laboratories (LabQAP). Spike recoveries outside of the required range were reported for Al, Ca, Cr, Fe, Li, K, Si, Na and Zr. Serial dilutions are performed by diluting a sample by five-fold. The results from the serial dilution may be found in the raw data and are signified by an "L" at the end of the sample number. For acceptable performance the percent difference between the serial dilution and the undiluted results must be ≤ 10 percent, one sample had a serial dilution greater than ten percent. This inaccuracy can be attributed to the low concentration of the analyte with respect to the detection limit. There did not appear to be any analytical anomalies or difficulties during the analyses of Ba or U.

Table 2. ICP Serial Dilution Raw Results for Tank 241-AN-106

Sample ID	Analyte	Sample Result Undiluted (ug/mL)	Serial Dilution Result (ug/mL)	% Difference
S96T003732	Al	43830.00	44750.00	2.1
S96T003732	Na	252600.00	259500.00	2.7
S96T003738	Na	239100.00	242800.00	1.5
S96T003739	Al	40860.00	41190.00	0.8
S96T003740	Al	40860.00	41250.00	0.9
S96T003740	Na	238800.00	243000.00	1.8
S96T003741	Al	40860.00	41190.00	0.8
S96T003741	Na	251000.00	255100.00	1.6
S96T003742	K	8678.00	8400.00	4.1
S96T003742	Al	42370.00	43690.00	3.1
S96T003742	Na	250900.00	260700.00	3.9
S96T003743	K	7098.00	7127.00	0.4
S96T003743	Al	45450.00	46130.00	1.5
S96T003743	Na	266500.00	271800.00	2.0
S96T003744	Na	213900.00	211500.00	1.1
S96T003744	Al	35000.00	36040.00	3.0
S96T003744	K	5513.00	5658.00	2.6
S96T003745	Na	236000.00	236000.00	0.0
S96T003745	Al	37500.00	37570.00	0.2
S96T003745	K	8056.00	5896.00	2.6
S96T003745	P	1240.00	1221.00	1.5
S96T003746	Na	240400.00	245900.00	2.3
S96T003746	Al	41290.00	42210.00	2.2
S96T003746	K	6192.00	6478.00	4.6
S96T003747	Al	41590.00	42400.00	1.9
S96T003747	Na	243700.00	249900.00	2.5
S96T003748	Na	241900.00	245700.00	1.6
S96T003748	K	6491.00	5425.00	16.4
S96T003749	Al	41440.00	41730.00	0.7
S96T003749	Na	240300.00	243200.00	1.2
S96T003750	Na	237500.00	240700.00	1.3
S96T003750	Al	40060.00	40320.00	0.6
S96T003750	K	6413.00	5795.00	9.6
S96T003761	K	8072.00	6128.00	0.9
S96T003761	Al	38700.00	39420.00	1.9
S96T003761	Na	234100.00	241500.00	3.2
S96T003804	Na	1081.00	1116.00	3.2
S96T003875	Al	35700.00	36480.00	7.8
S96T003875	Na	239200.00	245100.00	2.5
S96T003876	Na	217700.00	223100.00	2.5
S96T003876	Al	37350.00	38020.00	1.8
S96T003876	K	5810.00	5708.00	1.8
S96T003877	Na	252000.00	257400.00	2.1
S96T003877	Al	42260.00	43020.00	1.8
S96T003877	K	6682.00	6496.00	2.8
S96T003878	K	6290.00	6246.00	0.7
S96T003878	Na	236900.00	238300.00	0.6
S96T003879	K	6291.00	6757.00	7.4
S96T003879	Al	42930.00	43810.00	2.0
S96T003879	Na	252100.00	256500.00	1.7
S96T003980	Al	27420.00	27720.00	1.1
S96T003980	Na	223200.00	226200.00	1.3
S96T003988	Na	1013.00	1028.00	1.5
S96T004308	Na	258500.00	262800.00	1.7
S96T004309	K	6636.00	6603.00	0.5
S96T004309	Al	43190.00	43050.00	0.3
S96T004309	Na	254400.00	254700.00	0.1
S96T004333	Al	44310.00	44280.00	0.1
S96T004333	Na	268600.00	265000.00	0.2
S96T004419	Na	262200.00	265300.00	1.2
S96T004419	Al	44180.00	44190.00	0.0
S96T004420	Na	246300.00	251800.00	2.2
S96T004420	Al	41290.00	42670.00	1.9
S96T004420	K	6638.00	6861.00	3.4
S96T004421	Na	254700.00	254100.00	0.2
S96T004421	Al	42550.00	42280.00	0.6
S96T004421	K	6433.00	6795.00	5.6
S96T004433	Na	269100.00	275000.00	2.2
S96T004434	Na	233100.00	243000.00	4.2
S96T004434	Al	39400.00	37200.00	5.6
S96T004434	K	6066.00	6211.00	2.4
S96T004831	Na	254400.00	260100.00	2.2
S96T004831	K	6493.00	6470.00	0.4
S96T004831	Al	43390.00	44110.00	1.7

$$\% \text{ Difference} = (\text{ABS}(\text{Sample} - \text{Serial})) / \text{Sample} \times 100$$

Ion Chromatography (IC)

The IC analyses were performed on direct subsamples of liquid samples, as indicated by a blank in the A# column in Table 1. The solid subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1.

High RPDs (> 20%) were reported for several analytes and can be attributed to sample inhomogeneity. Spike recoveries outside of the required range (75% to 125%) were reported for Acetate, Cl, F, NO₂ and NO₃. The chemist noted that the spike failures were due to matrix interferences from organic acids. Bromide was detected in 34 of the 52 subsamples submitted for analysis and was an indication of hydrostatic head fluid intrusion into the samples. The standard recoveries for this analysis were within the required limits.

Total Inorganic/Organic Carbon Persulfate (TIC/TOC)

TIC/TOC by persulfate/coulometry analyses were performed on the core composite samples. None of the results exceeded the TOC notification limit of 30,000 µgC/g. Low concentrations of TIC and TOC were noted in some preparation blanks, however the levels of contamination are inconsequential when compared to the result of the sample and do not impact sample data quality. The standard recoveries, spike recoveries and RPDs for these analyses were within the required limits.

OH

The OH analyses were performed on direct liquid core composite subsamples and the field blank, as indicated by a blank in the A# column in Table 1. The solid core composite subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1. The standard recoveries and RPDs for this analysis were within the required limits.

Chromium VI (CrVI)

The CrVI analyses were performed on direct liquid core composite subsamples and the field blank, as indicated by a blank in the A# column in Table 1. The solid core composite subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1. The standard recoveries, spike recoveries and RPDs for this analysis were within the required limits.

Radionuclide Analyses**Total Alpha (AT)**

Total alpha activity analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in the A# column in Table 1. The results and the raw data for all subsamples are included in this revision. High RPDs (>20%) were reported for the solid core composite subsamples and can be attributed to sample inhomogeneity. No reruns were requested due to the low alpha activity in the samples. The standard recoveries and spike recoveries for this analysis were within the required limits.

Total Beta (TB)

Total beta activity analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. Low levels of beta activity were noted in some preparation blanks, however the levels of contamination are inconsequential when compared to the result of the sample and do not impact sample data quality. The standard recoveries, spike recoveries and RPDs for this analysis were within the required limits.

Strontium 90 (⁹⁰Sr)

⁹⁰Sr analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. The standard recoveries and RPDs for this analysis were within the required limits.

Gamma Energy Analysis (GEA)

GEA analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1.

Actual detection limits for GEA analytes are not currently available. The latest GEA software does not currently report a minimum detectable activity (MDA). If an analyte is reported as "less than", the value reported is the detection limit.

Low levels of cesium activity were noted in some preparation blanks, however the levels of contamination are inconsequential when compared to the result of the sample and do not impact sample data quality. The standard recoveries and RPDs for this analysis were within the required limits.

Total Uranium (U)

Uranium analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. The standard recoveries, spike recoveries and RPDs for this analysis were within the required limits.

Plutonium ($^{239/240}\text{Pu}$)

$^{239/240}\text{Pu}$ analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. The standard recoveries and RPDs for this analysis were within the required limits.

Iodine (^{129}I)

^{129}I analyses were performed on direct liquid core composite subsamples and the field blank, as indicated by a blank in the A# column in Table 1. The solid core composite subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1. The standard recoveries and RPDs for this analysis were within the required limits.

Technetium (^{99}Tc)

^{99}Tc analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. The standard recoveries and RPDs for this analysis were within the required limits.

Alpha Energy Analyses**Americium (^{241}Am)**

^{241}Am analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. High RPDs were reported for the solid core composite samples. This can be attributed to the heterogeneous nature of the samples. No rerun was requested due to the low americium activity in the samples. The standard recoveries for this analysis were within the required limits.

Curium ($^{243/244}\text{Cm}$)

$^{243/244}\text{Cm}$ analyses were performed on direct liquid core composite subsamples and the field blank. The solid core composite subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion preparation is indicated with an "F" in A# column in Table 1. The RPDs for this analysis were within the required limits.

Procedures

Table 3 lists the analytical procedures used for performing the sample analyses. Abbreviations for analyses are defined in the table notes.

Table 3. Analytical Procedures			
Analysis	Sample Portion	Preparation Procedure +	Analysis Procedure
DSC	Solid/Liquid	N/A	LA-514-114 Rev. C-1 LA-514-113 Rev. C-1
TGA	Solid/Liquid	N/A	LA-514-114 Rev. C-1 LA-560-112 Rev. B-1
Bulk Density	Solid	N/A	LO-160-103 Rev. B-0
Sp.G.	Liquid	N/A	LA-510-112 Rev. C-3
U	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-925-009 Rev. A-1
IC	Solid Liquid	LA-504-101 Rev. E-0 N/A	LA-533-105 Rev. D-1
ICP	Solid Liquid	LA-505-159 Rev. D-0 LA-549-141 Rev. F-0 N/A	LA-505-151 Rev. D-3 LA-505-161 Rev. B-1
OH-	Solid Liquid	LA-504-101 Rev. E-0 N/A	LA-211-102 Rev. C-0

Table 3. Analytical Procedures			
TICTOC	Solid/Liquid	N/A	LA-342-100 Rev. D-0
CR(VI)	Solid Liquid	LA-504-101 Rev. E-0 N/A	LA-265-101 Rev. B-0
TOC	Liquid	N/A	LA-344-105 Rev. D-1
^3H	Solid Liquid	LA-504-101 Rev. E-0 N/A	LA-218-114 Rev. B-0
^{129}I	Solid Liquid	LA-504-101 Rev. E-0 N/A	LA-378-103 Rev. C-0
^{99}Tc	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-438-101 Rev. D-2
^{237}Np	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-933-141 Rev. H-1
AT/TB	Solid Liquid	LA-549-101 Rev. F-0 N/A	LA-508-101 Rev. D-2
GEA	Solid Liquid	LA-549-101 Rev. F-0 N/A	LA-548-121 Rev. E-0
$^{241}\text{Am}/^{244}\text{Cm}$	Solid Liquid	LA-549-101 Rev. F-0 N/A	LA-953-103 Rev. B-0
^{90}Sr	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-220-101 Rev. D-1
^{239}Pu	Solid Liquid	LA-549-101 Rev. F-0 N/A	LA-943-128 Rev. B-0

Abbreviations:

N/A = not applicable (these are direct samples)
DSC = differential scanning calorimetry
TGA = thermogravimetric analysis
Sp.G. = specific gravity
% Water = percent water by gravimetric analysis
NH₃ = ammonia (by ion selective electrode (ISE))
OH = hydroxide
IC = ion chromatography
ICP = inductively coupled plasma
TOC = total organic carbon
TIC = total inorganic carbon
AT = total alpha
GEA = gamma energy analysis
¹³⁷Cs = cesium 137
⁹⁰Sr = strontium 90
²⁴¹Am = americium 241
^{239/240}Pu = plutonium 239/240

References

- Steen, F. H., 1997, *Tank 241-AN-105 Push Mode Core Sampling and Analysis Plan*, WHC-SD-WM-TSAP-027, Rev. 1-A, Westinghouse Hanford Company, Richland, WA 99352.
- Dukelow, G. T., J. W. Hunt, H. Babad, and J. E. Meacham, 1995, *Tank Safety Screening Data Quality Objective*, WHC-SD-WM-SP-004, Rev. 2, Westinghouse Hanford Company, Richland, WA 99352.
- Benar, C. J., 1995, *Flammable Gas Data Quality Objective*, WHC-SD-WM-DQO-004, Rev. 2, Westinghouse Hanford Company, Richland, WA 99352.

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

ATTACHMENT 1

AN-105 SAMPLE BREAKDOWN

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

AN-105
Core:152
Seg: 1
S96T003481

Attachment 1

15

IC-ANIONS
KCP
SFG

AN-105
Core:152
Seg: 2
S96T003482

Attachment 1

HNF-SD-WM-DP-199, REV. 1

AN-105
Core:152
Seg: 3
S96T003483

Attachment 1

AN-105
Core:152
Seg: 4
S96T003491

Attachment 1

AN-105
Core:152
Seg: 5
S96T003492

Attachment 1

AN-105
Core:152
Seg: 6
S96T003493

Attachment 1

AN-105
Core:152
Seg: 7
S96T003494

Attachment 1

AN-105
Core:152
Seg: 8
S96T003495

Attachment 1

AN-105
Core:152
Seg: 9
S96T003502

Attachment 1

AN-105
Core:152
Seg: 10
S96T003503

Attachment 1

AN-105
Core:152
Seg: 11
S96T003504

Attachment 1

AN-105
Core:152
Seg: 12
S96T003588

Attachment 1

AN-105
Core:152
Seg: 13
S96T003589

Attachment 1

AN-105
Core:152
Seg: 14
S96T003512

Attachment 1

AN-105
Core:152
Seg: 16
S96T003513

Attachment 1

AN-105
Core:152
Seg: 18
S96T003505

Attachment 1

AN-105
Core:152
Seg: 20
S96T003514

Attachment 1

HNF-SD-WM-DP-199, REV. 1

AN-105
Core:152
Seg: 22
S96T003590

Attachment 1

AN-105

Core:152

Composite:

Drainable Liquid

Attachment 1

AN-105
Core:152
Composite:
Segments 14-20

S96T004207

S96T004254

S96T004255

S96T004256

S96T004257

S96T004258

Composited:

Segment	Grams
14 LH	20.0
14 LH	20.0
16 LH	20.0
16 LH	20.0
18 LH	20.0
18 LH	20.0
20 LH	20.0
20 LH	20.0

Attachment 1

Composited:

Segment:	Grams:
22 UH	50.0
22 LH	50.0

AN-105
Core:152
Composite:
Segment 22

S96T004208

S96T004259

S96T004260

S96T004261

S96T004262

S96T004263

AN-105
Core:153
Seg: 1
S96T004384

Attachment 1

AN-105

Core:153

Seg: 2

S96T004154

Attachment 1

AN-105
Core:153
Seg: 3
S96T004385

Attachment 1

AN-105
Core:153
Seg: 5
S96T004326

Attachment 1

AN-105
Core:153
Seg: 6
S96T004327

Attachment 1

AN-105
Core:153
Seg: 7
S96T004323

Attachment 1

AN-105
Core:153
Seg: 8
S96T003664

Attachment 1

AN-105
Core:153
Seg: 9
S96T003637

Attachment 1

AN-105
Core:153
Seg: 10
S96T004155

Attachment 1

AN-105
Core:153
Seg: 11
S96T003638

Attachment 1

AN-105
Core:153
Seg: 12
S96T003639

Attachment 1

AN-105
Core:153
Seg: 13
S96T003665

Attachment 1

AN-105
Core:153
Seg: 14
S96T004324

Attachment 1

HNF-SD-WM-DP-199, REV. 1

AN-105
Core:153
Seg: 15
S96T004156

Attachment 1

AN-105
Core:153
Seg: 17
S96T004386

Attachment 1

AN-105
Core:153
Seg: 19
S96T004328

Attachment 1

AN-105
Core:153
Seg: 21
S96T003666

Attachment 1

HNF-SD-WM-DP-199, REV. 1

AN-105
Core:153
Seg: 22
S96T003630

Attachment 1

AN-105
Core:153
Seg: FB
S96T003663

Attachment 1

HNF-SD-WM-DP-199, REV. 1

Attachment 1

AN-105
Core:153
Seg: HHF
S96T003631

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

SAMPLE DATA SUMMARY

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

Table 1. Data Summary Table
AN-105

CORE NUMBER: 152
SEGMENT #: 1

SEGMENT PORTION: U Upper Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003714		Bulk Density of Sample	g/mL	n/a	n/a	1.520	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961003731		DSC Exotherm using Mettler	Joules/g	104.0	n/a	5.000	7.300	6.150	37.4	n/a	n/a	n/a
S961003731		DSC Exotherm DRY, Calculated	Joules/g Dry	n/a	n/a	8.890	13.00	10.95	37.6	n/a	n/a	n/a
S961003731		% Water by TGA using Mettler	%	100.9	n/a	46.48	40.81	43.75	13.4	n/a	n/a	n/a
S961003893		F Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<1.98e+02	<2.01e2	n/a	88.40	n/a	198.0	n/a
S961003893		F Aluminum - ICP-Fusion	ug/g	98.20	<5.00e-02	2.56e+04	2.44e+04	2.50e+04	4.80	89.00	988.0	n/a
S961003893		F Arsenic - ICP-Fusion	ug/g	103.4	<1.00e-01	<1.98e+03	<2.01e3	n/a	102.0	102.0	1.98e+03	n/a
S961003893		F Boron - ICP-Fusion	ug/g	100.0	<5.00e-02	<9.88e+02	<1.01e3	n/a	n/a	95.50	988.0	n/a
S961003893		F Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<9.88e+02	<1.01e3	n/a	n/a	94.50	988.0	n/a
S961003893		F Beryllium - ICP-Fusion	ug/g	103.6	<5.00e-03	<9.88e+02	<1.01e3	n/a	n/a	95.30	988.0	n/a
S961003893		F Bismuth - ICP-Fusion	ug/g	102.0	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	98.40	988.0	n/a
S961003893		F Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	100.0	1.98e+03	n/a
S961003893		F Cadmium - ICP-Fusion	ug/g	100.8	<5.00e-03	<9.88e+02	<1.01e3	n/a	n/a	97.40	988.0	n/a
S961003893		F Cerium - ICP-Fusion	ug/g	100.0	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	95.90	1.98e+03	n/a
S961003893		F Cobalt - ICP-Fusion	ug/g	100.8	<2.00e-02	<3.95e+02	<4.02e2	n/a	n/a	97.20	395.0	n/a
S961003893		F Chromium - ICP-Fusion	ug/g	101.4	<1.00e-02	7.21e+02	712.0	716.5	1.26	97.70	198.0	n/a
S961003893		F Copper - ICP-Fusion	ug/g	103.4	<1.00e-02	<1.98e+02	<2.01e2	n/a	n/a	98.10	198.0	n/a
S961003893		F Iron - ICP-Fusion	ug/g	101.0	<5.00e-01	<9.88e+02	<1.01e3	n/a	n/a	97.40	988.0	n/a
S961003893		F Potassium - ICP-Fusion	ug/g	96.80	<5.00e-01	8.26e+06	n/a	n/a	n/a	n/a	9.88e+03	n/a
S961003893		F Lanthanum - ICP-Fusion	ug/g	101.6	<5.00e-02	<9.88e+02	<1.01e3	n/a	n/a	97.60	988.0	n/a
S961003893		F Lithium - ICP-Fusion	ug/g	102.0	<1.00e-02	<1.98e+02	<2.01e2	n/a	n/a	95.90	198.0	n/a
S961003893		F Magnesium - ICP-Fusion	ug/g	97.20	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	91.50	1.98e+03	n/a
S961003893		F Manganese - ICP-Fusion	ug/g	98.20	<1.00e-02	<1.98e+02	<2.01e2	n/a	n/a	93.00	198.0	n/a
S961003893		F Molybdenum - ICP-Fusion	ug/g	100.4	<5.00e-02	<9.88e+02	<1.01e3	n/a	n/a	97.30	988.0	n/a
S961003893		F Sodium - ICP-Fusion	ug/g	98.80	1.700	2.38e+05	2.54e+05	2.46e+05	6.50	63.70	1.98e+03	n/a
S961003893		F Neodymium - ICP-Fusion	ug/g	101.8	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	97.00	1.98e+03	n/a
S961003893		F Nickel - ICP-Fusion	ug/g	100.4	1.24e-01	7.85e+02	2.92e+03	1.85e+03	115	97.50	395.0	n/a
S961003893		F Phosphorus - ICP-Fusion	ug/g	101.6	<2.00e-01	<3.95e+03	<4.02e3	n/a	n/a	106.0	3.95e+03	n/a
S961003893		F Lead - ICP-Fusion	ug/g	99.80	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	105.0	1.98e+03	n/a
S961003893		F Sulfur - ICP-Fusion	ug/g	99.00	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	93.30	1.98e+03	n/a
S961003893		F Antimony - ICP-Fusion	ug/g	95.60	<6.00e-02	<1.98e+03	<2.01e3	n/a	n/a	97.60	1.98e+03	n/a
S961003893		F Selenium - ICP-Fusion	ug/g	97.80	<1.00e-01	<9.88e+02	<1.01e3	n/a	n/a	97.60	988.0	n/a
S961003893		F Silicon - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.98e+02	<2.01e2	n/a	n/a	94.50	1.98e+03	n/a
S961003893		F Samarium - ICP-Fusion	ug/g	98.40	<1.00e-01	<1.98e+03	<2.01e3	n/a	n/a	94.90	198.0	n/a
S961003893		F Strontium - ICP-Fusion	ug/g	99.20	<1.00e-02	<1.98e+02	<2.01e2	n/a	n/a	93.70	198.0	n/a
S961003893		F Titanium - ICP-Fusion	ug/g	96.80	<1.00e-01	<3.95e+03	<4.02e3	n/a	n/a	94.40	3.95e+03	n/a
S961003893		F Thallium - ICP-Fusion	ug/g	99.00	<5.00e-01	<9.88e+02	<1.01e3	n/a	n/a	87.50	9.88e+03	n/a
S961003893		F Uranium - ICP-Fusion	ug/g	100.2	<5.00e-02	<9.88e+02	<1.01e3	n/a	n/a	96.50	988.0	n/a
S961003893		F Vanadium - ICP-Fusion	ug/g	101.2	<1.00e-02	7.04e+02	586.0	645.0	18.3	96.40	198.0	n/a
S961003893		F Zinc - ICP-Fusion	ug/g	98.40	<1.00e-02	<1.98e+02	<2.01e2	n/a	n/a	97.60	988.0	n/a
S961003893		F Zirconium - ICP-Fusion	ug/g	84.38	<1.44e-02	<1.42e+02	<2.28e+2	n/a	n/a	74.63	3.70e+02	5.00e-02
S961003906		A Silver - ICP-Acid Digest	ug/g	23.40	<1.00e-02	13.70	13.40	13.55	2.21	37.50	3.860	n/a

Sample#	R.A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003906	A	Aluminum - ICP-Acid Digest	ug/g	101.4	2.28e-01	2.51e+04	2.48e+04	2.50e+04	1.20	1.71e02	19.30	n/a
S96T003906	A	Arsenic - ICP-Acid Digest	ug/g	95.00	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	96.20	38.60	n/a
S96T003906	A	Boron - ICP-Acid Digest	ug/g	118.2	6.30e-01	97.40	111.0	104.2	13.1	93.60	19.30	n/a
S96T003906	A	Barium - ICP-Acid Digest	ug/g	92.80	< 5.00e-02	< 19.30	< 2.01e1	n/a	n/a	93.20	19.30	n/a
S96T003906	A	Beryllium - ICP-Acid Digest	ug/g	94.60	< 5.00e-03	< 1.930	< 2.01e0	n/a	n/a	94.60	1.930	n/a
S96T003906	A	Bismuth - ICP-Acid Digest	ug/g	92.40	< 1.00e-01	1.08e+02	< 4.03e1	n/a	n/a	94.20	38.60	n/a
S96T003906	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	< 38.60	< 4.03e1	145.0	51.0	95.20	38.60	n/a
S96T003906	A	Cadmium - ICP-Acid Digest	ug/g	90.80	< 5.00e-03	3.310	3.760	3.535	12.7	92.80	1.930	n/a
S96T003906	A	Cerium - ICP-Acid Digest	ug/g	95.00	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	96.20	38.60	n/a
S96T003906	A	Cobalt - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	7.710	< 8.06e0	n/a	n/a	94.80	7.720	n/a
S96T003906	A	Chromium - ICP-Acid Digest	ug/g	93.00	< 5.00e-02	7.58e+02	733.0	745.5	3.35	83.60	3.860	n/a
S96T003906	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	4.710	4.320	4.515	8.64	92.40	3.860	n/a
S96T003906	A	Iron - ICP-Acid Digest	ug/g	93.00	< 5.00e-02	44.80	32.80	38.80	30.9	90.40	19.30	n/a
S96T003906	A	Potassium - ICP-Acid Digest	ug/g	93.80	< 5.00e-01	3.99e+03	3.79e+03	3.88e+03	4.39	39.60	193.0	n/a
S96T003906	A	Lanthanum - ICP-Acid Digest	ug/g	95.00	< 5.00e-02	< 19.30	< 2.01e1	n/a	n/a	96.60	19.30	n/a
S96T003906	A	Lithium - ICP-Acid Digest	ug/g	96.60	< 1.00e-02	< 3.860	< 4.03e1	n/a	n/a	99.80	3.860	n/a
S96T003906	A	Magnesium - ICP-Acid Digest	ug/g	88.60	< 1.00e-02	< 3.860	< 4.03e1	n/a	n/a	87.60	38.60	n/a
S96T003906	A	Manganese - ICP-Acid Digest	ug/g	90.00	< 1.00e-02	< 3.860	< 4.03e1	n/a	n/a	90.00	3.860	n/a
S96T003906	A	Molybdenum - ICP-Acid Digest	ug/g	94.60	< 5.00e-02	5.780	< 4.03e1	57.20	2.16	94.80	19.30	n/a
S96T003906	A	Niobium - ICP-Acid Digest	ug/g	150.6	1.900	1.87e+05	1.78e+05	1.82e+05	4.93	-9.74e02	38.60	n/a
S96T003906	A	Nickel - ICP-Acid Digest	ug/g	94.00	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	95.20	38.60	n/a
S96T003906	A	Nickel - ICP-Acid Digest	ug/g	93.40	< 2.00e-02	12.40	16.30	14.35	25.8	95.00	7.730	n/a
S96T003906	A	Phosphorus - ICP-Acid Digest	ug/g	93.00	< 2.00e-01	1.33e+02	939.0	1.13e+02	34.2	66.20	71.60	n/a
S96T003906	A	Selenium - ICP-Acid Digest	ug/g	89.00	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	93.80	38.60	n/a
S96T003906	A	Sulfur - ICP-Acid Digest	ug/g	93.00	1.19e-01	1.58e+03	1.50e+03	1.54e+03	5.79	78.00	39.60	n/a
S96T003906	A	Tantalum - ICP-Acid Digest	ug/g	88.40	< 5.00e-02	< 25.70	< 2.01e1	n/a	n/a	90.80	25.70	n/a
S96T003906	A	Tellurium - ICP-Acid Digest	ug/g	87.40	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	91.40	38.60	n/a
S96T003906	A	Thallium - ICP-Acid Digest	ug/g	244.0	5.31e-01	2.19e+02	201.0	210.0	8.57	76.20	19.30	n/a
S96T003906	A	Titanium - ICP-Acid Digest	ug/g	92.20	< 1.00e-01	< 38.60	< 4.03e1	n/a	n/a	93.20	38.60	n/a
S96T003906	A	Titanium - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	< 3.860	< 4.03e1	n/a	n/a	92.40	3.860	n/a
S96T003906	A	Titanium - ICP-Acid Digest	ug/g	88.00	< 2.00e-01	< 77.10	< 8.06e1	n/a	n/a	86.00	71.60	n/a
S96T003906	A	Thallium - ICP-Acid Digest	ug/g	90.70	< 5.00e-01	< 1.93e+02	< 2.01e2	n/a	n/a	92.30	19.30	n/a
S96T003906	A	Vanadium - ICP-Acid Digest	ug/g	91.80	< 5.00e-02	< 19.30	< 2.01e1	n/a	n/a	92.40	19.30	n/a
S96T003906	A	Zinc - ICP-Acid Digest	ug/g	88.80	1.40e-02	< 32.30	< 4.03e1	31.70	3.78	91.40	3.860	n/a
S96T003906	A	Zinc - ICP-Acid Digest	ug/g	93.20	< 1.00e-02	< 3.860	< 4.03e1	n/a	n/a	94.20	3.860	n/a
S96T003912	W	Acetate by IC-Dionex 4000/4500	ug/g	100.8	< 2.00e-02	7.57e+02	< 1.59e2	1.30e+03	9.78	97.50	147.4	n/a
S96T003912	W	Formate by IC-Dionex 4000/4500	ug/g	100.8	< 2.00e-02	1.36e+03	1.24e03	1.30e+03	6.78	97.50	147.4	n/a
S96T003912	W	Fluoride by IC-Dionex 4000/4500	ug/g	96.95	< 3.00e-02	2.89e+02	5.5270	5.49e02	4.92	102.7	23.07	n/a
S96T003912	W	Chloride by IC-Dionex 4000/4500	ug/g	99.49	< 1.70e-02	7.92e+02	7.33e02	7.29e+02	4.39	107.0	436.8	n/a
S96T003912	W	Nitrite by IC-Dionex 4000/4500	ug/g	97.95	< 1.20e-01	5.14e+02	5.12e02	n/a	n/a	97.96	514.3	n/a
S96T003912	W	Bromide by Ion Chromatography	ug/g	97.98	< 1.60e-01	3.62e+02	1.58e02	1.53e+02	3.86	98.82	51.3	n/a
S96T003915	W	Nitrate by IC-Dionex 4000/4500	ug/g	105.2	< 1.60e-01	3.62e+02	3.35e02	3.58e+02	0.59	102.4	462.3	n/a
S96T003915	W	Phosphate by IC-Dionex 4000/4500	ug/g	97.94	< 1.30e-01	3.62e+02	3.43e02	3.44e+02	0.58	98.26	354.8	n/a
S96T003915	W	Sulfate by IC-Dionex 4000/4500	ug/g	98.49	< 1.00e-01	4.36e+03	4.39e+03	4.47e+03	3.80	99.62	428.6	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Drainable Liquid: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003732		DSC Exotherm using Mettler	Joules/g	98.77	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003732		DSC Exotherm Dry Calculated	Joules/g Dry	98.77	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003732		% water by IGA using Mettler	Sp. G.	99.73	n/a	4.97	51.74	50.66	1.06	n/a	1.00e-03	n/a
S961003732		D Silver-ICP-Acid Dil.	ug/mL	100.6	< 1.00e-02	18.00	17.20	17.60	4.28	n/a	4.010	n/a
S961003732		D Aluminum-ICP-Acid Dil.	ug/mL	98.80	< 5.00e-02	4.36e+04	4.28e+04	4.33e+04	2.31	92.60	4.010	n/a
S961003732		D Arsenic-ICP-Acid Dil.	ug/mL	104.2	< 1.00e-01	40.10	44.01e1	43.50	9.50	8.10e1	20.10	n/a
S961003732		D Boron-ICP-Acid Dil.	ug/mL	100.2	< 5.00e-02	72.40	70.40	71.40	2.80	95.50	40.10	n/a
S961003732		D Barium-ICP-Acid Dil.	ug/mL	103.4	< 5.00e-02	< 20.10	< 2.01e1	n/a	n/a	97.80	20.10	n/a
S961003732		D Beryllium-ICP-Acid Dil.	ug/mL	103.4	< 5.00e-02	< 2.000	< 2.00e0	n/a	n/a	101.0	2.000	n/a
S961003732		D Bismuth-ICP-Acid Dil.	ug/mL	103.2	< 1.00e-01	< 40.10	44.01e1	n/a	n/a	103.0	40.10	n/a
S961003732		D Calcium-ICP-Acid Dil.	ug/mL	101.6	< 1.00e-01	< 40.10	44.01e1	n/a	n/a	103.0	40.10	n/a
S961003732		D Cadmium-ICP-Acid Dil.	ug/mL	101.6	< 5.00e-03	< 2.000	< 2.00e0	n/a	n/a	99.60	2.000	n/a
S961003732		D Cerium-ICP-Acid Dil.	ug/mL	101.4	< 1.00e-01	< 40.10	44.01e1	n/a	n/a	99.30	40.10	n/a
S961003732		D Cobalt-ICP-Acid Dil.	ug/mL	102.0	< 2.00e-02	< 8.020	8.02e0	n/a	n/a	98.10	8.020	n/a
S961003732		D Chromium-ICP-Acid Dil.	ug/mL	104.0	< 1.00e-02	< 4.010	< 4.01e0	229.5	3.92	98.10	4.010	n/a
S961003732		D Copper-ICP-Acid Dil.	ug/mL	101.6	< 5.00e-02	< 20.10	< 2.01e1	n/a	n/a	101.0	20.10	n/a
S961003732		D Potassium-ICP-Acid Dil.	ug/mL	96.20	< 5.00e-01	6.89e+03	6.66e+03	6.78e+03	3.39	89.70	200.0	n/a
S961003732		D Lanthanum-ICP-Acid Dil.	ug/mL	102.6	< 5.00e-02	< 20.10	< 2.01e1	n/a	n/a	101.0	20.10	n/a
S961003732		D Lithium-ICP-Acid Dil.	ug/mL	97.00	< 1.00e-01	< 40.10	44.01e0	n/a	n/a	105.0	40.10	n/a
S961003732		D Magnesium-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-02	< 4.010	< 4.01e0	n/a	n/a	95.00	40.10	n/a
S961003732		D Manganese-ICP-Acid Dil.	ug/mL	101.6	< 5.00e-02	< 1.01e+02	< 1.01e1	n/a	n/a	95.00	40.10	n/a
S961003732		D Molybdenum-ICP-Acid Dil.	ug/mL	98.80	< 1.00e-01	< 2.53e+05	2.48e+05	99.25	3.53	99.80	20.10	n/a
S961003732		D Neodymium-ICP-Acid Dil.	ug/mL	103.2	< 1.00e-01	< 40.10	44.01e1	2.50e+05	2.81	-9.370e2	40.10	n/a
S961003732		D Nickel-ICP-Acid Dil.	ug/mL	101.8	< 2.00e-02	< 8.020	8.02e0	n/a	n/a	101.0	40.10	n/a
S961003732		D Phosphorus-ICP-Acid Dil.	ug/mL	99.20	< 2.00e-01	4.08e+02	353.0	380.5	14.5	91.50	80.20	n/a
S961003732		D Lead-ICP-Acid Dil.	ug/mL	101.4	< 1.00e-01	< 51.50	51.50	52.75	4.74	100.0	40.10	n/a
S961003732		D Sulfur-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-01	< 1.01e+03	980.0	999.0	3.02	95.40	40.10	n/a
S961003732		D Antimony-ICP-Acid Dil.	ug/mL	96.00	< 5.00e-02	< 24.10	< 2.41e1	n/a	n/a	95.40	24.10	n/a
S961003732		D Selenium-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-01	< 40.10	44.01e1	n/a	n/a	103.0	40.10	n/a
S961003732		D Silicon-ICP-Acid Dil.	ug/mL	99.00	< 5.00e-02	2.24e+02	220.0	222.0	1.80	96.40	20.10	n/a
S961003732		D Strontium-ICP-Acid Dil.	ug/mL	100.2	< 1.00e-02	< 40.10	< 4.01e1	n/a	n/a	97.90	40.10	n/a
S961003732		D Tantalum-ICP-Acid Dil.	ug/mL	98.40	< 1.00e-02	< 4.010	< 4.01e0	n/a	n/a	98.30	4.010	n/a
S961003732		D Thallium-ICP-Acid Dil.	ug/mL	97.20	< 2.00e-02	80.20	80.02e1	n/a	n/a	96.90	4.010	n/a
S961003732		D Uranium-ICP-Acid Dil.	ug/mL	97.10	< 5.00e-01	< 2.00e+02	< 2.00e2	n/a	n/a	98.00	200.0	n/a
S961003732		D Vanadium-ICP-Acid Dil.	ug/mL	102.2	< 5.00e-02	< 20.10	< 2.01e1	n/a	n/a	99.20	20.10	n/a
S961003732		D Zinc-ICP-Acid Dil.	ug/mL	99.40	< 1.00e-02	23.50	23.50	23.50	0.00	98.40	4.010	n/a
S961003732		D Zirconium-ICP-Acid Dil.	ug/mL	98.30	< 2.00e-02	4.710	4.710	4.740	1.27	98.00	4.010	n/a
S961003732		Acetate by IC-Dioxex 4000/4500	ug/mL	99.30	< 2.00e-02	7.60e+02	680.0	724.2	12.2	n/a	22.22	n/a
S961003732		Formate by Ion Chromatograph	ug/mL	101.7	< 2.00e-02	1.74e+03	1.66e+03	1.70e+03	4.71	n/a	23.33	n/a
S961003732		Fluoride-IC-Dioxex 4000/4500	ug/mL	108.3	< 1.30e-02	< 35.83	35.83e1	n/a	n/a	n/a	53.83	n/a
S961003732		Chloride-IC-Dioxex 4000/4500	ug/mL	99.20	< 1.00e-02	9.86e+03	9.38e+03	9.62e+03	4.99	n/a	70.40	n/a
S961003732		Nitrite-IC - Dioxex 4000/4500	ug/mL	99.27	< 1.07e-01	1.20e+05	1.25e+05	1.27e+05	2.37	n/a	1.08e+03	n/a
S961003732		Bromide by Ion Chromatograph	ug/mL	100.0	< 1.28e-01	< 5.22e+02	656.0	n/a	n/a	n/a	321.8	n/a
S961003732		Nitrate by IC-Dioxex 4000/4500	ug/mL	100.0	< 1.47e-01	1.59e+05	1.58e+05	1.58e+05	0.63	n/a	1.41e+03	n/a

Sample#	R/A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003732		Phosphate-IC-Dionex 4000/4500	ug/mL	102.2	<1.19e-01	1.24e+03	494.0	865.5	86.0	n/a	692.1	n/a
S96T003732		Sulfate by IC-Dionex 4000/4500	ug/mL	101.1	<1.36e-01	1.04e+03	1.18e+03	1.11e+03	12.6	n/a	592.8	n/a
S96T003732		Oxalate by IC-Dionex 4000/4500	ug/mL	104.7	<1.05e-01	<4.35e+02	<4.35e+02	n/a	n/a	n/a	434.8	n/a
S96T003732		Alpha in Liquid Samples	uCi/mL	112.5	<1.73e-02	<2.28e-02	2.23e-02	n/a	n/a	103.1	2.70e-02	1.55E+02

CORE NUMBER: 152
SEGMENT #: 2

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	RA# Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Sok Rec %	Det Limit	Count Err%
S961003739	DSC Exotherm using Mettler	Joules/g	98.77	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003739	DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003739	Specific Gravity	Sp. G.	98.60	n/a	1.408	1.410	1.409	0.14	n/a	1.00e-03	n/a
S961003739	% Water by TGA using Mettler	%	99.73	n/a	50.20	50.57	50.39	0.73	n/a	n/a	n/a
S961003739	D Silver-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	16.90	17.00	16.95	0.59	91.10	4.010	n/a
S961003739	D Arsenic-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	4.09e+04	3.91e+04	4.00e+04	4.50	-8.12e+02	20.10	n/a
S961003739	D Aluminum-ICP-Acid Dil.	ug/mL	104.2	<1.00e-01	40.10	<4.01e1	73.25	1.77	89.30	40.10	n/a
S961003739	D Boron-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	73.90	72.60	73.25	1.77	96.30	20.10	n/a
S961003739	D Barium-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	98.80	20.10	n/a
S961003739	D Beryllium-ICP-Acid Dil.	ug/mL	103.4	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	103.0	2.000	n/a
S961003739	D Bismuth-ICP-Acid Dil.	ug/mL	103.2	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	101.0	40.10	n/a
S961003739	D Calcium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	103.0	40.10	n/a
S961003739	D Cadmium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	100.0	2.000	n/a
S961003739	D Cesium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	101.0	40.10	n/a
S961003739	D Cobalt-ICP-Acid Dil.	ug/mL	102.4	<2.00e-02	< 8.020	-8.02e0	n/a	n/a	99.90	8.020	n/a
S961003739	D Chromium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	2.01e+02	198.0	199.5	1.50	99.50	4.010	n/a
S961003739	D Copper-ICP-Acid Dil.	ug/mL	104.0	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	102.0	4.010	n/a
S961003739	D Iron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	99.90	20.10	n/a
S961003739	D Potassium-ICP-Acid Dil.	ug/mL	96.20	<5.00e-01	6.42e+03	6.50e+03	6.46e+03	1.24	95.80	200.0	n/a
S961003739	D Lanthanum-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	103.0	20.10	n/a
S961003739	D Lithium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	5.130	5.180	5.155	0.97	105.0	4.010	n/a
S961003739	D Magnesium-ICP-Acid Dil.	ug/mL	97.00	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	93.50	40.10	n/a
S961003739	D Manganese-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	95.50	4.010	n/a
S961003739	D Molybdenum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	94.70	95.00	94.85	0.32	101.0	20.10	n/a
S961003739	D Sodium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	2.39e+05	2.40e+05	2.40e+05	0.42	65.30	40.10	n/a
S961003739	D Neodymium-ICP-Acid Dil.	ug/mL	103.2	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	103.0	40.10	n/a
S961003739	D Nickel-ICP-Acid Dil.	ug/mL	101.8	<2.00e-02	< 8.020	-8.02e0	n/a	n/a	99.10	8.020	n/a
S961003739	D Phosphorus-ICP-Acid Dil.	ug/mL	99.20	<2.00e-01	3.46e+02	337.0	341.5	2.64	92.30	80.20	n/a
S961003739	D Lead-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	64.00	48.60	64.30	9.94	101.0	40.10	n/a
S961003739	D Sulfur-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	9.30e+02	711.0	820.5	26.7	32.20	40.10	n/a
S961003739	D Antimony-ICP-Acid Dil.	ug/mL	98.60	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	95.80	20.10	n/a
S961003739	D Selenium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	102.0	40.10	n/a
S961003739	D Silicon-ICP-Acid Dil.	ug/mL	98.00	<5.00e-02	1.89e+02	185.0	187.0	2.14	95.10	20.10	n/a
S961003739	D Strontium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	99.10	40.10	n/a
S961003739	D Tantalum-ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	97.60	4.010	n/a
S961003739	D Tellurium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	91.90	80.20	n/a
S961003739	D Thallium-ICP-Acid Dil.	ug/mL	97.20	<5.00e-01	< 80.20	-8.02e1	n/a	n/a	98.00	200.0	n/a
S961003739	D Uranium-ICP-Acid Dil.	ug/mL	101.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	99.70	20.10	n/a
S961003739	D Vanadium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	18.60	22.00	20.30	16.7	102.0	4.010	n/a
S961003739	D Zinc-ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	4.260	<4.01e0	n/a	n/a	99.30	2.222	n/a
S961003739	D Zirconium-ICP-Acid Dil.	ug/mL	99.30	<2.00e-02	7.48e+02	727.0	737.2	2.85	154.0	20.10	n/a
S961003739	Formate by Ion Chromatograph	ug/mL	101.7	<2.10e-02	1.65e+03	1.71e+03	1.68e+03	3.57	115.0	23.35	n/a

13-Jan-1997 10:51:38
A-0002-3

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count	Err%
S961003739		Fluoride-IC-Dionex 4000/4500	ug/mL	95.59	<1.70e-02	4.95e-02	466.0	480.4	6.04	n/a	66.96	n/a
S961003739		Chloride-IC-Dionex 4000/4500	ug/mL	102.3	<1.70e-02	9.24e-03	9.29e+03	9.27e+03	0.54	n/a	87.57	n/a
S961003739		Nitrite-IC-Dionex 4000/4500	ug/mL	98.53	<1.07e-01	1.20e-03	1.21e+05	1.20e+05	0.83	n/a	551.2	n/a
S961003739		Bromide by Ion Chromatograph	ug/mL	104.6	<1.26e-01	1.03e-03	1.06e+03	1.04e+03	2.87	n/a	649.0	n/a
S961003739		Nitrate by IC-Dionex 4000/4500	ug/mL	103.0	<1.72e-01	1.60e-02	1.61e+05	1.61e+05	0.62	n/a	721.1	n/a
S961003739		Phosphate-IC-Dionex 4000/4500	ug/mL	99.08	<1.19e-01	<6.12e-02	1.01e+03	n/a	n/a	n/a	612.5	n/a
S961003739		Sulfate by IC-Dionex 4000/4500	ug/mL	102.9	<1.36e-01	9.78e-02	1.32e+03	1.15e+03	29.8	n/a	700.0	n/a
S961003739		Oxalate by IC-Dionex 4000/4500	ug/mL	101.9	<1.05e-01	<5.41e-02	<5.41e-02	n/a	n/a	n/a	540.9	n/a
S961003739		Alpha in Liquid Samples	ug/mL	112.5	<1.73e-02	<2.47e-02	<2.72e-2	n/a	n/a	n/a	2.70e-02	2.77e-02

CORE NUMBER: 152
SEGMENT #: 3

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003740		DSC Exotherm using Mettler	Joules/g	98.77	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003740		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003740		Specific Gravity	SP-G	98.49	n/a	1.408	1.410	1.408	0.28	n/a	1.00e-03	n/a
S961003740		% Water by TSA on Perkin Elmer	%	99.27	n/a	50.53	49.29	50.06	1.48	n/a	n/a	n/a
S961003740		D Silver-ICP-Acid Dil.	ug/mL	100.6	< 1.00e-02	16.70	16.80	16.75	1.79	90.90	4.010	n/a
S961003740		D Aluminum-ICP-Acid Dil.	ug/mL	98.80	< 1.00e-02	4.09e+04	3.78e+04	3.94e+04	7.88	-6.130e1	20.10	n/a
S961003740		D Arsenic-ICP-Acid Dil.	ug/mL	104.2	< 1.00e-02	78.10	44.01e1	n/a	n/a	89.90	40.10	n/a
S961003740		D Boron-ICP-Acid Dil.	ug/mL	100.2	< 1.00e-02	78.10	74.70	76.55	4.65	95.60	20.10	n/a
S961003740		D Barium-ICP-Acid Dil.	ug/mL	100.8	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	97.00	20.10	n/a
S961003740		D Beryllium-ICP-Acid Dil.	ug/mL	103.4	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	101.0	2.000	n/a
S961003740		D Bismuth-ICP-Acid Dil.	ug/mL	103.2	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	102.0	40.10	n/a
S961003740		D Calcium-ICP-Acid Dil.	ug/mL	101.6	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	105.0	40.10	n/a
S961003740		D Cadmium-ICP-Acid Dil.	ug/mL	101.6	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	99.90	2.000	n/a
S961003740		D Cerium-ICP-Acid Dil.	ug/mL	101.4	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	99.80	40.10	n/a
S961003740		D Cobalt-ICP-Acid Dil.	ug/mL	102.0	< 1.00e-02	< 8.02e+02	< 8.02e+02	198.5	3.55	99.60	8.020	n/a
S961003740		D Chromium-ICP-Acid Dil.	ug/mL	102.2	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	99.90	4.010	n/a
S961003740		D Copper-ICP-Acid Dil.	ug/mL	104.0	< 1.00e-02	< 4.010	< 4.010	n/a	n/a	99.80	4.010	n/a
S961003740		D Iron-ICP-Acid Dil.	ug/mL	101.6	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	99.80	4.010	n/a
S961003740		D Potassium-ICP-Acid Dil.	ug/mL	96.20	< 1.00e-02	6.42e+03	6.40e+03	6.41e+03	0.51	86.50	20.10	n/a
S961003740		D Lanthanum-ICP-Acid Dil.	ug/mL	102.6	< 1.00e-02	< 2.00	< 2.00	n/a	n/a	101.0	20.10	n/a
S961003740		D Lithium-ICP-Acid Dil.	ug/mL	102.2	< 1.00e-02	< 4.010	< 4.010	n/a	n/a	99.80	4.010	n/a
S961003740		D Magnesium-ICP-Acid Dil.	ug/mL	97.00	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	99.80	4.010	n/a
S961003740		D Manganese-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-02	< 4.010	< 4.010	n/a	n/a	99.80	4.010	n/a
S961003740		D Molybdenum-ICP-Acid Dil.	ug/mL	101.6	< 1.00e-02	< 95.50	92.80	94.15	2.87	95.60	40.10	n/a
S961003740		D Sodium-ICP-Acid Dil.	ug/mL	98.80	< 1.00e-02	2.39e+05	2.34e+05	2.36e+05	2.11	-9.85e3	20.10	n/a
S961003740		D Neodymium-ICP-Acid Dil.	ug/mL	103.2	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	99.80	40.10	n/a
S961003740		D Nickel-ICP-Acid Dil.	ug/mL	101.8	< 1.00e-02	< 8.02e+02	< 8.02e+02	338.0	2.37	99.10	8.020	n/a
S961003740		D Phosphorus-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-02	< 43.80	43.80	45.90	9.15	100.0	40.10	n/a
S961003740		D Lead-ICP-Acid Dil.	ug/mL	101.4	< 1.00e-02	48.00	43.80	45.90	9.15	100.0	40.10	n/a
S961003740		D Sulphur-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-02	9.51e+02	697.0	824.0	30.8	96.10	40.10	n/a
S961003740		D Antimony-ICP-Acid Dil.	ug/mL	96.00	< 1.00e-02	< 24.10	< 24.10	n/a	n/a	95.40	24.10	n/a
S961003740		D Selenium-ICP-Acid Dil.	ug/mL	98.80	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	103.0	40.10	n/a
S961003740		D Silicon-ICP-Acid Dil.	ug/mL	96.00	< 1.00e-02	2.18e+02	205.0	211.5	6.15	96.10	20.10	n/a
S961003740		D Samarium-ICP-Acid Dil.	ug/mL	99.00	< 1.00e-02	< 40.10	< 40.10	n/a	n/a	97.50	40.10	n/a
S961003740		D Strontium-ICP-Acid Dil.	ug/mL	100.2	< 1.00e-02	< 4.010	< 4.010	n/a	n/a	94.40	4.010	n/a
S961003740		D Titanium-ICP-Acid Dil.	ug/mL	98.40	< 1.00e-02	< 80.20	< 80.20	n/a	n/a	97.20	80.20	n/a
S961003740		D Thallium-ICP-Acid Dil.	ug/mL	97.20	< 1.00e-02	< 80.20	< 80.20	n/a	n/a	97.50	20.10	n/a
S961003740		D Uranium-ICP-Acid Dil.	ug/mL	101.2	< 1.00e-02	< 22.10	< 22.10	n/a	n/a	99.40	20.10	n/a
S961003740		D Vanadium-ICP-Acid Dil.	ug/mL	102.2	< 1.00e-02	< 20.10	< 20.10	n/a	n/a	102.0	20.10	n/a
S961003740		D Zinc-ICP-Acid Dil.	ug/mL	94.36	< 1.00e-02	4.540	4.190	4.365	8.02	98.00	4.010	n/a
S961003740		Acetate by IC-Dioxen 4000/4500	ug/mL	94.36	< 1.00e-02	6.87e+02	716.0	701.5	8.02	93.63	424.0	n/a
S961003740		Formate by Ion Chromatograph	ug/mL	101.7	< 1.00e-02	1.54e+03	1.53e+03	1.54e+03	0.05	112.5	424.0	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003740		Fluoride-IC-Dionex 4000/4500	ug/mL	101.2	<1.30e-02	5.07e+02	503.0	504.9	0.79	60.51	66.96	N/A
S96T003740		Chloride-IC-Dionex 4000/4500	ug/mL	107.1	<1.70e-02	1.00e+04	1.02e+04	1.01e+04	1.98	103.0	87.57	N/A
S96T003740		Nitrite-IC - Dionex 4000/4500	ug/mL	99.26	<1.07e-01	1.29e+05	1.28e+05	1.29e+05	0.78	98.34	551.2	N/A
S96T003740		Bromide by Ion Chromatograph	ug/mL	103.1	<1.26e-01	1.23e+03	1.16e+03	1.19e+03	5.86	99.66	649.0	N/A
S96T003740		Nitrate by IC-Dionex 4000/4500	ug/mL	106.6	2.36e-01	1.72e+05	1.74e+05	1.73e+05	1.16	108.2	721.1	N/A
S96T003740		Phosphate-IC-Dionex 4000/4500	ug/mL	99.82	<1.19e-01	<6.12e+02	1.10e+03	N/A	N/A	102.2	612.5	N/A
S96T003740		Sulfate by IC-Dionex 4000/4500	ug/mL	103.6	<1.36e-01	1.27e+03	1.10e+03	1.14e+03	23.8	101.6	700.0	N/A
S96T003740		Oxalate by IC-Dionex 4000/4500	ug/mL	103.6	<1.08e-01	<5.41e+02	<5.41e+02	N/A	N/A	106.2	540.9	N/A
S96T003740		Alpha In Liquid Samples	ug/mL	112.5	<1.73e-02	<1.65e-02	4.33e-02	N/A	N/A	N/A	2.70e-02	2.77e-02

Table 1. Data Summary Table
AN-105

CORE NUMBER: 152
SEGMENT #: 4

SEGMENT PORTION: Drainable Liquid

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Error
S96T003741		DSC Exotherm using Mettler	Joules/g	98.77	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003741		DSC Exotherm	Joules/g DRY		n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003741		Specific Gravity	Sp G.	98.60	n/a	1.384	1.401	1.397	0.50	n/a	1.00e-03	n/a
S96T003741		% Water by TGA on Perkin Elmer		99.27	n/a	50.12	50.83	50.47	1.41	n/a	1.00e-03	n/a
S96T003741		Silver-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	17.90	16.70	17.30	6.94	n/a	4.0	n/a
S96T003741		Aluminum-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	4.09e+04	3.9e+04	4.02e+04	3.74	n/a	20.10	n/a
S96T003741		Arsenic-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	<4.0	<4.0	<4.0	89.70	n/a	40.10	n/a
S96T003741		Boron-ICP-Acid Dil.	ug/mL	104.2	<5.00e-02	<77.00	71.20	74.10	7.83	n/a	20.10	n/a
S96T003741		Barium-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	20.10	n/a
S96T003741		Beryllium-ICP-Acid Dil.	ug/mL	103.4	<5.00e-02	<2.000	<2.00e0	n/a	n/a	n/a	20.10	n/a
S96T003741		Bismuth-ICP-Acid Dil.	ug/mL	103.2	<1.00e-02	<4.0	<4.0	n/a	n/a	n/a	40.10	n/a
S96T003741		Calcium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	<4.0	<4.0	n/a	n/a	n/a	40.10	n/a
S96T003741		Cadmium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	<2.000	<2.00e0	n/a	n/a	n/a	20.10	n/a
S96T003741		Carbon-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	<4.0	<4.0	n/a	n/a	n/a	20.10	n/a
S96T003741		Chromium-ICP-Acid Dil.	ug/mL	102.0	<2.00e-02	<8.020	<8.02e0	209.0	4.78	n/a	4.01e1	n/a
S96T003741		Cobalt-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	<2.16e+02	204.0	n/a	n/a	n/a	8.01e1	n/a
S96T003741		Copper-ICP-Acid Dil.	ug/mL	104.0	<1.00e-02	<4.01e	<4.01e1	n/a	n/a	n/a	20.10	n/a
S96T003741		Iron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	20.10	n/a
S96T003741		Potassium-ICP-Acid Dil.	ug/mL	96.20	<5.00e-02	6.75e+03	6.41e+03	6.58e+03	5.17	n/a	83.40	n/a
S96T003741		Lanthanum-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	20.10	n/a
S96T003741		Lithium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	<5.240	<5.24e0	5.250	0.00	n/a	40.10	n/a
S96T003741		Magnesium-ICP-Acid Dil.	ug/mL	97.00	<1.00e-02	<4.0	<4.0	n/a	n/a	n/a	40.10	n/a
S96T003741		Manganese-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	<4.01e	<4.01e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Molybdenum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	<4.01e	<4.01e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Sodium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	98.60	93.70	96.15	5.18	n/a	20.10	n/a
S96T003741		Nickel-ICP-Acid Dil.	ug/mL	103.2	<1.00e-02	2.51e+05	2.36e+05	2.42e+05	7.01	n/a	40.10	n/a
S96T003741		Neodymium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	<4.0	<4.0	n/a	n/a	n/a	40.10	n/a
S96T003741		Phosphorus-ICP-Acid Dil.	ug/mL	101.8	<2.00e-02	<8.020	<8.02e0	324.0	5.98	n/a	8.020	n/a
S96T003741		Lead-ICP-Acid Dil.	ug/mL	99.20	<2.00e-02	3.44e+02	3.24e0	334.0	5.98	n/a	40.10	n/a
S96T003741		Sulfur-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	51.85	46.90	49.35	9.93	n/a	40.10	n/a
S96T003741		Antimony-ICP-Acid Dil.	ug/mL	96.60	<5.00e-02	<24.10	<2.41e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Selenium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	8.10e+02	758.0	784.0	6.63	n/a	40.10	n/a
S96T003741		Silicon-ICP-Acid Dil.	ug/mL	96.00	<5.00e-02	2.00e+02	198.0	202.0	3.96	n/a	20.10	n/a
S96T003741		Strontium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	<40.16	<4.01e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Thallium-ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	<4.01e	<4.01e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Titanium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	<4.01e	<4.01e1	n/a	n/a	n/a	40.10	n/a
S96T003741		Thallium-ICP-Acid Dil.	ug/mL	97.20	<2.00e-02	<80.20	<8.02e1	n/a	n/a	n/a	200.10	n/a
S96T003741		Vanadium-ICP-Acid Dil.	ug/mL	97.10	<5.00e-02	<20.12	<2.01e1	n/a	n/a	n/a	20.10	n/a
S96T003741		Uranium-ICP-Acid Dil.	ug/mL	101.2	<5.00e-02	<18.60	<1.86e1	17.60	11.4	n/a	4.010	n/a
S96T003741		Zinc-ICP-Acid Dil.	ug/mL	102.6	<1.00e-02	4.360	4.320	4.340	0.88	n/a	4.010	n/a
S96T003741		Zirconium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	7.26e+02	4.24e2	4.340	98.60	n/a	424.0	n/a
S96T003741		Acetate by IC-Dioxex 4000/4500	ug/mL	94.96	<2.00e-01	1.26e+02	<4.24e03	n/a	n/a	n/a	424.0	n/a
S96T003741		Formate by Ion Chromatograph	ug/mL	101.7	<2.00e-01	1.34e+03	1.54e+03	0.00	n/a	n/a	424.0	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R#	AnalYte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003741		Fluoride-IC-Dionex 4000/4500	ug/mL	101.2	<1.30e-02	5.08e+02	477.0	492.8	6.29	n/a	64.96	n/a
S96T003741		Chloride-IC-Dionex 4000/4500	ug/mL	107.1	<1.70e-02	9.92e+03	1.02e+04	1.01e+04	2.78	n/a	87.57	n/a
S96T003741		Nitrite-IC - Dionex 4000/4500	ug/mL	99.26	<1.07e-01	1.26e+05	1.27e+05	1.26e+05	0.79	n/a	551.2	n/a
S96T003741		Bromide by Ion Chromatograph	ug/mL	103.1	<1.26e-01	8.54e+02	<6.49e2	n/a	n/a	n/a	449.2	n/a
S96T003741		Nitrate by IC-Dionex 4000/4500	ug/mL	106.6	<1.38e-01	1.64e+05	1.65e+05	1.64e+05	0.61	n/a	721.0	n/a
S96T003741		Phosphate-IC-Dionex 4000/4500	ug/mL	99.82	<1.19e-01	1.21e+03	<6.12e2	n/a	n/a	n/a	52.5	n/a
S96T003741		Sulfate by IC-Dionex 4000/4500	ug/mL	103.6	<1.34e-01	1.08e+03	1.13e+03	1.10e+03	4.52	n/a	700.0	n/a
S96T003741		Oxalate by IC-Dionex 4000/4500	ug/mL	103.6	<1.08e-01	<5.41e+02	<5.41e2	n/a	n/a	n/a	540.9	n/a
S96T003741		Alpha in Liquid Samples	ug/mL	112.5	<1.79e-02	<2.88e-02	<2.47e-2	n/a	n/a	n/a	2.70e-02	5.00e+02

Table 1. Data Summary Table
AN-105

CORE NUMBER: 152
SEGMENT #: 5

SEGMENT POSITION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
594T0037/2		DSC Exotherm using Mettler	Joules/g	97.01	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
594T0037/2		DSC Exotherm, Calculated	Joules/g	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
594T0037/2		Specific Gravity	Sp. G.	98.60	n/a	1.416	1.406	1.411	0.71	n/a	1.00e-03	n/a	n/a
594T0037/2		3 Meter, by GA using Mettler	Sp. G.	99.76	n/a	50.50	50.85	50.67	0.69	n/a	88.90	4.010	n/a
594T0037/2		D Silver-TP-Acid Dil.	ug/mL	100.6	< 1.00e-02	18.10	17.40	17.75	3.94	n/a	87.20	40.10	n/a
594T0037/2		D Lithium-TP-Acid Dil.	ug/mL	98.80	< 5.00e-02	4.24e+04	3.97e+04	4.10e+04	6.58	n/a	87.20	40.10	n/a
594T0037/2		D Boron-TP-Acid Dil.	ug/mL	104.2	< 1.00e-01	73.10	68.10	70.60	7.08	n/a	96.50	20.10	n/a
594T0037/2		D Sodium-TP-Acid Dil.	ug/mL	100.8	< 5.00e-02	2.00	< 2.01e1	n/a	n/a	n/a	101.0	2.000	n/a
594T0037/2		D Barium-TP-Acid Dil.	ug/mL	103.4	< 1.00e-01	40.10	< 4.01e1	n/a	n/a	n/a	103.0	40.10	n/a
594T0037/2		D Beryllium-TP-Acid Dil.	ug/mL	101.6	< 1.00e-01	40.10	< 4.01e1	n/a	n/a	n/a	105.0	40.10	n/a
594T0037/2		D Calcium-TP-Acid Dil.	ug/mL	103.2	< 1.00e-01	40.10	< 4.01e1	n/a	n/a	n/a	99.40	2.000	n/a
594T0037/2		D Cadmium-TP-Acid Dil.	ug/mL	101.4	< 1.00e-01	40.10	< 4.01e1	n/a	n/a	n/a	98.80	40.10	n/a
594T0037/2		D Cobalt-TP-Acid Dil.	ug/mL	102.0	< 1.00e-02	8.020	< 8.02e0	n/a	n/a	n/a	99.50	8.020	n/a
594T0037/2		D Chromium-TP-Acid Dil.	ug/mL	102.2	< 1.00e-02	2.11e+02	202.0	206.3	4.36	n/a	96.80	4.010	n/a
594T0037/2		D Copper-TP-Acid Dil.	ug/mL	104.0	< 1.00e-02	4.010	< 4.01e0	n/a	n/a	n/a	100.0	4.010	n/a
594T0037/2		D Lead-TP-Acid Dil.	ug/mL	101.6	< 5.00e-02	20.10	< 2.01e1	n/a	n/a	n/a	99.10	20.10	n/a
594T0037/2		D Potassium-TP-Acid Dil.	ug/mL	96.20	< 5.00e-01	6.68e+03	6.55e+03	6.62e+03	1.97	n/a	51.40	200.0	n/a
594T0037/2		D Lithium-TP-Acid Dil.	ug/mL	102.6	< 5.00e-02	20.10	< 2.01e1	n/a	n/a	n/a	100.0	20.10	n/a
594T0037/2		D Magnesium-TP-Acid Dil.	ug/mL	97.00	< 1.00e-01	4.590	< 4.01e1	n/a	n/a	n/a	103.0	4.010	n/a
594T0037/2		D Manganese-TP-Acid Dil.	ug/mL	99.20	< 1.00e-02	40.10	< 4.01e1	n/a	n/a	n/a	94.70	40.10	n/a
594T0037/2		D Nickel-TP-Acid Dil.	ug/mL	101.6	< 5.00e-02	4.010	< 4.01e0	n/a	n/a	n/a	94.80	4.010	n/a
594T0037/2		D Sodium-TP-Acid Dil.	ug/mL	98.80	< 1.00e-01	2.51e+05	2.39e+05	98.30	3.46	n/a	98.80	20.10	n/a
594T0037/2		D Neodymium-TP-Acid Dil.	ug/mL	103.2	< 1.00e-01	40.10	< 4.01e1	2.45e+05	4.90	n/a	101.0	40.10	n/a
594T0037/2		D Nickel-TP-Acid Dil.	ug/mL	101.8	< 2.00e-02	8.020	< 8.02e0	n/a	n/a	n/a	99.70	8.020	n/a
594T0037/2		D Phosphorus-TP-Acid Dil.	ug/mL	99.20	< 2.00e-01	3.60e+02	330.0	345.0	8.70	n/a	92.60	80.20	n/a
594T0037/2		D Lead-TP-Acid Dil.	ug/mL	101.4	< 1.00e-01	51.60	49.30	50.45	4.56	n/a	99.70	40.10	n/a
594T0037/2		D Sul Fur-TP-Acid Dil.	ug/mL	96.60	< 5.00e-02	24.10	< 2.41e1	839.5	16.8	n/a	52.90	40.10	n/a
594T0037/2		D Antimony-TP-Acid Dil.	ug/mL	98.00	< 1.00e-01	9.10e+02	769.0	n/a	n/a	n/a	95.20	24.10	n/a
594T0037/2		D Selenium-TP-Acid Dil.	ug/mL	96.00	< 5.00e-02	2.72e+02	228.0	250.0	17.6	n/a	102.0	40.10	n/a
594T0037/2		D Silicon-TP-Acid Dil.	ug/mL	99.00	< 1.00e-01	40.10	< 4.01e1	n/a	n/a	n/a	83.80	20.10	n/a
594T0037/2		D Strontium-TP-Acid Dil.	ug/mL	100.2	< 1.00e-02	4.010	< 4.01e1	n/a	n/a	n/a	97.10	40.10	n/a
594T0037/2		D Thallium-TP-Acid Dil.	ug/mL	98.40	< 1.00e-02	4.010	< 4.01e0	n/a	n/a	n/a	96.90	4.010	n/a
594T0037/2		D Vanadium-TP-Acid Dil.	ug/mL	97.20	< 2.00e-01	80.20	< 8.02e1	n/a	n/a	n/a	91.20	80.20	n/a
594T0037/2		D Uranium-TP-Acid Dil.	ug/mL	101.2	< 5.00e-01	2.00e+02	2.00e2	n/a	n/a	n/a	97.00	200.0	n/a
594T0037/2		D Zirconium-TP-Acid Dil.	ug/mL	102.2	< 1.00e-02	20.10	< 2.01e1	n/a	n/a	n/a	98.80	20.10	n/a
594T0037/2		D Zinc-TP-Acid Dil.	ug/mL	102.2	< 1.00e-02	25.90	20.10	23.00	25.2	n/a	100.0	4.010	n/a
594T0037/2		D Arsenic by IC-Dioxin	ug/mL	99.60	< 1.00e-02	5.100	4.920	5.010	3.59	n/a	97.30	4.010	n/a
594T0037/2		Acetate by IC-Dioxin	ug/mL	106.5	< 2.00e-01	4.88e+02	< 4.24e2	n/a	n/a	n/a	n/a	424.0	n/a
596T0037/2		Formate by Ion Chromatograph	ug/mL	104.2	< 2.00e-01	1.54e+03	1.58e+03	1.56e+03	2.56	n/a	n/a	424.0	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003742		Fluoride-IC-Dionex 4000/4500	ug/mL	99.32	<1.50e-02	<1.33e+02	<1.33e2	n/a	n/a	n/a	132.6	n/a
S96T003742		Chloride-IC-Dionex 4000/4500	ug/mL	95.82	<1.70e-02	2.87e+03	2.90e+03	2.89e+03	1.04	n/a	173.4	n/a
S96T003742		Nitrite-IC - Dionex 4000/4500	ug/mL	96.69	<1.07e-01	3.22e+04	3.09e+04	3.15e+04	4.12	n/a	1.09e+03	n/a
S96T003742		Bromide by Ion Chromatograph	ug/mL	99.15	<1.26e-01	<1.28e+03	<1.29e3	n/a	n/a	n/a	1.28e+03	n/a
S96T003742		Nitrate by IC-Dionex 4000/4500	ug/mL	97.14	1.76e-01	3.81e+04	3.71e+04	3.76e+04	2.66	n/a	1.43e+03	n/a
S96T003742		Phosphate-IC-Dionex 4000/4500	ug/mL	96.70	<1.19e-01	<1.21e+03	<1.21e3	n/a	n/a	n/a	1.21e+03	n/a
S96T003742		Sulfate by IC-Dionex 4000/4500	ug/mL	97.78	<1.36e-01	<1.39e+03	<1.39e3	n/a	n/a	n/a	1.39e+03	n/a
S96T003742		Oxalate by IC-Dionex 4000/4500	ug/mL	100.2	<1.05e-01	<1.07e+03	<1.07e3	n/a	n/a	n/a	1.07e+03	n/a
S96T003742		Alpha in Liquid Samples	uCi/mL	112.5	<1.73e-02	<1.65e-02	<2.28E-2	n/a	n/a	n/a	2.70e-02	5.00E-02

CORE NUMBER: 152
SEGMENT #: 6

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Error
S941003743	D	DSC Exotherm using Mettler	Joules/g	97.01	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S941003743	D	DSC Exotherm DRY	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S941003743	D	Specific Gravity	Sp. G.	98.60	n/a	1.399	1.408	1.403	0.64	1.00e-03	n/a
S941003743	D	% Water by TGA using Mettler	%	99.76	<1.00e-02	50.07	50.40	50.23	0.66	n/a	n/a
S941003743	D	Aluminum-ICP-Acid Dil.	ug/mL	97.80	<5.00e-02	4.55e+02	4.79e+02	4.67e+02	5.14	-5.87e-02	20.10
S941003743	D	Arsenic-ICP-Acid Dil.	ug/mL	103.0	<1.00e-01	<4.01e-01	<4.01e-01	<4.01e-01	1.01	91.98	4.01e-03
S941003743	D	Boron-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	77.40	83.80	80.70	7.68	94.40	20.10
S941003743	D	Beryllium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-03	<2.00e-02	<2.00e-02	<2.00e-02	10.2	102.0	2.00e-03
S941003743	D	Bismuth-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	<4.01e-01	<4.01e-01	<4.01e-01	10.8	108.0	4.01e-03
S941003743	D	Calcium-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	<4.01e-01	<4.01e-01	<4.01e-01	17.8	178.0	4.01e-03
S941003743	D	Cadmium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-03	<2.00e-02	<2.00e-02	<2.00e-02	17.8	178.0	2.00e-03
S941003743	D	Cerium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-01	<4.01e-01	<4.01e-01	<4.01e-01	17.8	178.0	2.00e-03
S941003743	D	Cobalt-ICP-Acid Dil.	ug/mL	100.8	<2.00e-02	8.020	<8.02e-02	26.0	7.35	94.30	8.02e-03
S941003743	D	Chromium-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	2.36e+02	25.0	26.0	7.35	94.30	8.02e-03
S941003743	D	Copper-ICP-Acid Dil.	ug/mL	103.4	<5.00e-02	5.050	5.220	5.135	3.31	101.0	4.01e-03
S941003743	D	Iron-ICP-Acid Dil.	ug/mL	101.8	<5.00e-02	<2.01e-01	<2.01e-01	<2.01e-01	10.2	102.0	2.00e-03
S941003743	D	Lithium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-01	7.10e+03	7.52e+03	7.31e+03	5.75	-2.38e-01	20.10
S941003743	D	Lead-ICP-Acid Dil.	ug/mL	100.6	<5.00e-02	<2.01e-01	<2.01e-01	<2.01e-01	10.2	102.0	2.00e-03
S941003743	D	Magnesium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	5.370	6.560	5.965	19.9	105.0	2.00e-03
S941003743	D	Manganese-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	<4.01e-01	<4.01e-01	<4.01e-01	17.8	178.0	4.01e-03
S941003743	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	1.03e+02	109.0	106.6	5.74	100.0	20.10
S941003743	D	Nickel-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	2.66e+05	2.77e+05	2.72e+05	4.05	-4.23e-03	40.10
S941003743	D	Selenium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	<4.01e-01	<4.01e-01	<4.01e-01	10.2	102.0	4.01e-03
S941003743	D	Silver-ICP-Acid Dil.	ug/mL	100.6	<2.00e-02	10.40	10.30	10.45	2.87	99.60	8.02e-03
S941003743	D	Sodium-ICP-Acid Dil.	ug/mL	99.40	<2.00e-01	4.06e+02	4.68e+02	437.0	13.7	99.50	80.20
S941003743	D	Sulfur-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	55.30	61.00	57.10	13.7	99.50	40.10
S941003743	D	Titanium-ICP-Acid Dil.	ug/mL	95.60	<5.00e-02	<2.01e-01	<2.01e-01	<2.01e-01	10.2	102.0	2.00e-03
S941003743	D	Vanadium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-02	2.53e+02	271.0	262.0	6.87	94.10	20.10
S941003743	D	Zinc-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	<4.01e-01	<4.01e-01	<4.01e-01	n/a	95.10	20.10
S941003743	D	Strontium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	<4.01e-01	<4.01e-01	<4.01e-01	n/a	99.20	40.10
S941003743	D	Thallium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	<4.01e-01	<4.01e-01	<4.01e-01	n/a	97.60	4.01e-03
S941003743	D	Uranium-ICP-Acid Dil.	ug/mL	95.00	<2.00e-01	<80.20	<8.02e-01	<8.02e-01	n/a	95.80	80.20
S941003743	D	Vanadium-ICP-Acid Dil.	ug/mL	95.00	<5.00e-01	<2.00e-02	<2.00e-02	<2.00e-02	n/a	100.0	200.0
S941003743	D	Zinc-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	<2.01e-01	<2.01e-01	<2.01e-01	n/a	100.0	20.10
S941003743	D	Zinc-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	27.30	30.50	29.30	3.56	104.0	4.01e-03
S941003743	D	Reactive by IC-Dioxane 4000/4500	ug/mL	104.2	<2.00e-01	<4.24e-02	<4.24e-02	<4.24e-02	23.0	99.10	4.01e-03
S941003743	D	Formate by Ion Chromatograph	ug/mL	104.2	<2.00e-01	1.73e+03	1.87e+03	1.70e+03	3.53	106.7	424.0

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R. A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003743		Fluoride-IC-Dionex 4000/4500	ug/mL	99.32	<1.30e-02	<1.33e+02	<1.33e2	n/a	n/a	169.2	132.6	n/a
S96T003743		Chloride-IC-Dionex 4000/4500	ug/mL	95.82	<1.70e-02	2.54e+04	2.58e+04	2.56e+04	1.56	164.6	173.4	n/a
S96T003743		Nitrite-IC - Dionex 4000/4500	ug/mL	96.69	<1.07e-01	3.29e+05	3.29e+05	3.29e+05	0.00	198.5	1.09e+03	n/a
S96T003743		Bromide by Ion Chromatograph	ug/mL	99.15	<1.26e-01	1.67e+03	1.66e+03	1.66e+03	0.60	95.76	1.28e+03	n/a
S96T003743		Nitrate by IC-Dionex 4000/4500	ug/mL	97.14	1.76e-01	4.19e+05	4.19e+05	4.19e+05	0.00	195.3	1.43e+03	n/a
S96T003743		Phosphate-IC-Dionex 4000/4500	ug/mL	96.70	<1.19e-01	1.95e+03	1.81e+03	1.88e+03	7.45	101.7	1.21e+03	n/a
S96T003743		Sulfate by IC-Dionex 4000/4500	ug/mL	97.78	<1.36e-01	2.90e+03	3.76e+03	3.33e+03	25.8	102.4	1.39e+03	n/a
S96T003743		Oxalate by IC-Dionex 4000/4500	ug/mL	100.2	<1.05e-01	<1.07e+03	<1.07e3	n/a	n/a	106.2	1.07e+03	n/a
S96T003743		Alpha In Liquid Samples	uCi/mL	100.0	<3.04e-02	<2.44e-02	<4.23E-2	n/a	n/a	86.35	3.70e-02	5.00e+02

CORE NUMBER: 152
SEGMENT #: 7

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003744	D	BSC Exotherm using Mettler	Joules/g	97.72	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003744	D	BSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003744	D	Specific Gravity	Sp. G.	99.53	n/a	1.437	1.432	1.434	0.35	n/a	1.00e-03	n/a
S961003744	D	% Water by TGA using Mettler	%	99.78	n/a	50.91	50.37	50.64	1.07	n/a	n/a	n/a
S961003744	D	Silver-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	3.60e+04	3.98e+04	3.79e+04	10.0	1.61e+03	20.10	n/a
S961003744	D	Aluminum-ICP-Acid Dil.	ug/mL	97.80	<1.00e-02	15.50	17.00	16.25	9.23	88.90	4.010	n/a
S961003744	D	Arsenic-ICP-Acid Dil.	ug/mL	103.0	<1.00e-01	<4.010	<4.010	<4.010	n/a	88.70	40.10	n/a
S961003744	D	Boron-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	63.60	71.60	67.60	11.8	101.0	20.10	n/a
S961003744	D	Barium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	<2.010	<2.00e0	n/a	n/a	98.70	20.10	n/a
S961003744	D	Beryllium-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	<4.010	<4.010	n/a	n/a	103.0	2.000	n/a
S961003744	D	Bismuth-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	<4.010	<4.010	n/a	n/a	106.0	40.10	n/a
S961003744	D	Calcium-ICP-Acid Dil.	ug/mL	99.00	<5.00e-03	<2.000	<2.00e0	n/a	n/a	98.90	2.000	n/a
S961003744	D	Cadmium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-01	<4.010	<4.010	n/a	n/a	99.30	40.10	n/a
S961003744	D	Cerium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-01	<4.010	<4.010	n/a	n/a	99.60	8.020	n/a
S961003744	D	Cobalt-ICP-Acid Dil.	ug/mL	100.6	<2.00e-02	8.020	<8.020	n/a	n/a	107.0	4.010	n/a
S961003744	D	Chromium-ICP-Acid Dil.	ug/mL	103.4	<1.00e-02	1.81e+02	199.0	190.0	9.47	107.0	4.010	n/a
S961003744	D	Copper-ICP-Acid Dil.	ug/mL	101.8	<1.00e-02	<4.010	<4.010	n/a	n/a	103.0	20.10	n/a
S961003744	D	Iron-ICP-Acid Dil.	ug/mL	99.60	<5.00e-01	5.51e+03	6.05e+03	5.78e+03	9.34	351.0	200.0	n/a
S961003744	D	Potassium-ICP-Acid Dil.	ug/mL	100.6	<5.00e-02	<2.010	<2.010	n/a	n/a	101.0	20.10	n/a
S961003744	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	11.00	<4.010	n/a	n/a	104.0	4.010	n/a
S961003744	D	Lithium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	<4.010	<4.010	n/a	n/a	95.20	40.10	n/a
S961003744	D	Magnesium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	<4.010	<4.010	n/a	n/a	95.10	4.010	n/a
S961003744	D	Manganese-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	82.50	92.00	87.25	10.9	104.0	20.10	n/a
S961003744	D	Molybdenum-ICP-Acid Dil.	ug/mL	98.00	<5.00e-01	2.14e+05	2.35e+05	2.24e+05	9.35	8.84e+03	40.10	n/a
S961003744	D	Sodium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	<4.010	<4.010	n/a	n/a	102.0	40.10	n/a
S961003744	D	Niobium-ICP-Acid Dil.	ug/mL	100.6	<2.00e-02	3.83e+02	371.0	377.0	3.18	107.0	8.020	n/a
S961003744	D	Nickel-ICP-Acid Dil.	ug/mL	106.4	<2.00e-01	43.20	45.10	44.15	4.30	101.0	40.10	n/a
S961003744	D	Phosphorus-ICP-Acid Dil.	ug/mL	99.40	<1.00e-01	8.50e+02	948.0	899.0	10.9	139.0	40.10	n/a
S961003744	D	Lead-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	<4.010	<4.010	n/a	n/a	95.00	24.10	n/a
S961003744	D	Sulfur-ICP-Acid Dil.	ug/mL	95.60	<6.00e-02	<2.010	<2.010	n/a	n/a	104.0	40.10	n/a
S961003744	D	Antimony-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	1.95e+02	215.0	205.0	9.76	109.0	20.10	n/a
S961003744	D	Selenium-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	<4.010	<4.010	n/a	n/a	98.10	40.10	n/a
S961003744	D	Silicon-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	<4.010	<4.010	n/a	n/a	98.70	4.010	n/a
S961003744	D	Samarium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	<4.010	<4.010	n/a	n/a	96.80	4.010	n/a
S961003744	D	Strontium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	<4.010	<4.010	n/a	n/a	95.80	80.20	n/a
S961003744	D	Titanium-ICP-Acid Dil.	ug/mL	96.00	<2.00e-01	80.20	<8.020	n/a	n/a	97.00	200.0	n/a
S961003744	D	Thallium-ICP-Acid Dil.	ug/mL	95.40	<5.00e-01	<2.00e+02	<2.00e2	n/a	n/a	97.00	20.10	n/a
S961003744	D	Uranium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	<2.010	<2.010	n/a	n/a	11.9	4.010	n/a
S961003744	D	Vanadium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	32.50	36.60	34.55	11.9	103.0	4.010	n/a
S961003744	D	Zinc-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	<4.010	<4.010	n/a	n/a	99.40	4.010	n/a
S961003744	D	Zirconium-ICP-Acid Dil.	ug/mL	94.36	<2.00e-01	6.72e+02	570.0	620.8	16.4	n/a	424.0	n/a
S961003744	D	Acetate by IC-Dioxen 4000/4500	ug/mL	101.7	<2.00e-01	1.60e+03	1.45e+03	1.55e+03	9.84	n/a	n/a	n/a
S961003744	D	Formate by Ion Chromatograph	ug/mL	101.7	<2.00e-01	1.60e+03	1.45e+03	1.55e+03	9.84	n/a	n/a	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003744		Fluoride-IC-Dionex 4000/4500	ug/mL	101.0	<1.30e-02	4.38e+02	429.0	433.3	2.08	38.81	53.83	n/a
S96T003744		Chloride-IC-Dionex 4000/4500	ug/mL	111.1	<1.70e-02	1.01e+04	1.00e+04	1.00e+04	1.00	114.6	70.40	n/a
S96T003744		Nitrite-IC - Dionex 4000/4500	ug/mL	99.45	<1.07e-01	1.21e+03	1.22e+03	1.21e+03	0.82	105.7	1.09e+03	n/a
S96T003744		Bromide by Ion Chromatograph	ug/mL	109.2	<1.26e-01	1.06e+03	1.08e+03	1.07e+03	1.87	98.47	521.8	n/a
S96T003744		Nitrate by IC-Dionex 4000/4500	ug/mL	101.2	<1.40e-01	1.69e+05	1.69e+05	1.69e+05	0.00	112.0	1.43e+03	n/a
S96T003744		Phosphate-IC-Dionex 4000/4500	ug/mL	110.1	<1.19e-01	7.48e+02	668.0	707.8	11.3	104.2	492.4	n/a
S96T003744		Sulfate by IC-Dionex 4000/4500	ug/mL	111.4	<1.36e-01	1.09e+03	1.16e+03	1.12e+03	6.22	105.4	562.8	n/a
S96T003744		Oxalate by IC-Dionex 4000/4500	ug/mL	101.7	<1.05e-01	<4.35e+02	494.0	n/a	n/a	101.3	434.8	n/a
S96T003744		Alpha In Liquid Samples	uCi/mL	100.0	<3.04e-02	1.76e-02	<2.15E-2	n/a	n/a	n/a	3.70e-02	1.26E+02

VH-C-SC-WM-DF-19E, REV. 1

VH-C-SC-WM-DF-19E, REV. 1

CORE NUMBER: 152
SEGMENT #: 8

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961003745		DSC Exotherm using Mettler	Joules/g	97.72	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S961003745		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S961003745		Specific Gravity	SP-G.	99.53	n/a	1.347	1.353	1.350	0.44	n/a	1.00e-03	n/a	n/a
S961003745		% Water by TGA using Mettler	%	99.78	n/a	50.42	52.16	51.29	3.39	n/a	n/a	n/a	n/a
S961003745		D Silver-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	17.40	18.65	18.25	13.4	86.50	4.010	n/a	n/a
S961003745		D Aluminum-ICP-Acid Dil.	ug/mL	97.80	<5.00e-02	3.75e+04	19.90	19.90	21.9	1.15e+03	20.10	n/a	n/a
S961003745		D Arsenic-ICP-Acid Dil.	ug/mL	103.0	<1.00e-01	< 40.10	<4.01e1	4.21e+04	21.9	1.15e+03	20.10	n/a	n/a
S961003745		D Boron-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 40.10	<4.01e1	83.60	21.1	90.30	40.10	n/a	n/a
S961003745		D Barium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	< 20.10	<2.01e1	97.20	20.10	97.20	20.10	n/a	n/a
S961003745		D Beryllium-ICP-Acid Dil.	ug/mL	102.8	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	101.0	2.000	n/a	n/a
S961003745		D Bismuth-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	104.0	40.10	n/a	n/a
S961003745		D Calcium-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.20	2.000	n/a	n/a
S961003745		D Cadmium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	98.20	2.000	n/a	n/a
S961003745		D Cerium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.20	40.10	n/a	n/a
S961003745		D Cobalt-ICP-Acid Dil.	ug/mL	100.8	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	98.30	8.020	n/a	n/a
S961003745		D Chromium-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	1.95e+02	238.0	216.5	19.9	104.0	4.010	n/a	n/a
S961003745		D Copper-ICP-Acid Dil.	ug/mL	103.4	<1.00e-02	4.690	<4.01e0	n/a	n/a	99.60	4.010	n/a	n/a
S961003745		D Iron-ICP-Acid Dil.	ug/mL	101.8	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	102.0	20.10	n/a	n/a
S961003745		D Potassium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-01	6.06e+03	6.44e+03	6.44e+03	11.9	153.0	200.0	n/a	n/a
S961003745		D Lanthanum-ICP-Acid Dil.	ug/mL	100.6	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	99.80	20.10	n/a	n/a
S961003745		D Lithium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	27.30	5.780	16.54	130	98.30	4.010	n/a	n/a
S961003745		D Magnesium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	93.70	40.10	n/a	n/a
S961003745		D Manganese-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	94.40	4.010	n/a	n/a
S961003745		D Molybdenum-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	85.30	108.0	96.65	23.5	103.0	20.10	n/a	n/a
S961003745		D Sodium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	2.36e+05	2.71e+05	2.54e+05	13.8	2.77e+03	40.10	n/a	n/a
S961003745		D Neodymium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	100.0	40.10	n/a	n/a
S961003745		D Nickel-ICP-Acid Dil.	ug/mL	100.6	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	98.50	8.020	n/a	n/a
S961003745		D Phosphorus-ICP-Acid Dil.	ug/mL	106.4	<2.00e-02	1.24e+03	414.0	827.0	99.9	8.200e+1	80.20	n/a	n/a
S961003745		D Lead-ICP-Acid Dil.	ug/mL	99.40	<1.00e-01	47.10	49.70	48.40	5.37	98.90	40.10	n/a	n/a
S961003745		D Sulfur-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	8.16e+02	969.0	892.5	17.1	146.0	40.10	n/a	n/a
S961003745		D Antimony-ICP-Acid Dil.	ug/mL	95.60	<6.00e-02	< 24.10	<2.41e1	n/a	n/a	93.50	24.10	n/a	n/a
S961003745		D Selenium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	105.0	40.10	n/a	n/a
S961003745		D Silicon-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	2.65e+02	292.0	278.5	9.69	98.20	20.10	n/a	n/a
S961003745		D Samarium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	96.40	40.10	n/a	n/a
S961003745		D Strontium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	97.20	4.010	n/a	n/a
S961003745		D Titanium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	95.50	4.010	n/a	n/a
S961003745		D Thallium-ICP-Acid Dil.	ug/mL	96.00	<2.00e-01	< 80.20	<8.02e1	n/a	n/a	91.40	80.20	n/a	n/a
S961003745		D Uranium-ICP-Acid Dil.	ug/mL	95.40	<5.00e-02	<2.00e+02	<2.00e2	n/a	n/a	95.50	200.0	n/a	n/a
S961003745		D Vanadium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	98.70	20.10	n/a	n/a
S961003745		D Zinc-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	< 26.60	35.00	30.80	27.3	101.0	4.010	n/a	n/a
S961003745		D Zirconium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	< 40.10	<4.01e1	n/a	n/a	98.10	4.010	n/a	n/a
S961003745		Acetate by IC-Dionex 4000/4500	ug/mL	90.32	<2.00e-01	4.47e+02	459.0	452.8	2.65	83.06	424.0	n/a	n/a
S961003745		Formate by Ion Chromatograph	ug/mL	98.33	<2.00e-01	1.48e+03	1.48e+03	1.48e+03	0.00	106.7	424.0	n/a	n/a

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003745			Fluoride-IC-Dionex 4000/4500	ug/mL	100.7	<1.30e-02	< 53.83	<5.38e1	n/a	n/a	150.7	53.83	n/a
S96T003745			Chloride-IC-Dionex 4000/4500	ug/mL	99.49	2.40e-02	9.73e+03	9.72e+03	9.72e+03	0.10	95.06	70.40	n/a
S96T003745			Nitrite-IC - Dionex 4000/4500	ug/mL	97.98	1.24e-01	1.26e+05	1.25e+05	1.26e+05	0.80	97.43	1.09e+03	n/a
S96T003745			Bromide by Ion Chromatograph	ug/mL	101.2	<1.26e-01	7.73e+02	801.0	787.0	3.56	94.91	521.8	n/a
S96T003745			Nitrate by IC-Dionex 4000/4500	ug/mL	98.82	2.15e-01	1.63e+05	1.61e+05	1.62e+05	1.23	96.13	1.43e+03	n/a
S96T003745			Phosphate-IC-Dionex 4000/4500	ug/mL	99.63	<1.19e-01	6.42e+02	658.0	650.0	2.46	99.63	492.4	n/a
S96T003745			Sulfate by IC-Dionex 4000/4500	ug/mL	98.73	<1.36e-01	9.51e+02	937.0	943.9	1.48	99.68	562.8	n/a
S96T003745			Oxalate by IC-Dionex 4000/4500	ug/mL	102.6	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	103.8	434.8	n/a
S96T003745			Alpha In Liquid Samples	ug/mL	100.0	<3.04e-02	<2.74e-2	<2.74E-2	n/a	n/a	n/a	3.70e-02	5.00E+02

CORE NUMBER: 152
SEGMENT #: 9

SEGMENT PORTION: Drainable Liquid

Table 1. Data Summary Table
AN-105

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003746		DSC Exotherm using Mettler	Joules/g	94.20	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003746		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003746		Specific Gravity	Sp. G.	99.53	n/a	1.408	1.400	1.404	0.57	n/a	1.00e-03	n/a
S961003746	D	% Water by TGA on Perkin Elmer	%	99.31	n/a	50.80	50.03	50.41	1.53	n/a	n/a	n/a
S961003746	D	Silver-ICP-Acid Dill.	ug/mL	99.20	<1.00e-02	18.00	18.00	18.00	88.10	n/a	4.010	n/a
S961003746	D	Aluminum-ICP-Acid Dill.	ug/mL	97.80	<5.00e-02	4.19e+04	4.25e+04	4.19e+04	2.86	n/a	20.10	n/a
S961003746	D	Arsenic-ICP-Acid Dill.	ug/mL	103.0	<1.00e-01	<40.10	<4.01e1	<40.10	97.20	n/a	40.10	n/a
S961003746	D	Boron-ICP-Acid Dill.	ug/mL	99.80	<5.00e-02	75.80	75.80	75.80	0.00	n/a	20.10	n/a
S961003746	D	Barium-ICP-Acid Dill.	ug/mL	99.60	<5.00e-02	<20.10	<2.01e1	<20.10	99.90	n/a	20.10	n/a
S961003746	D	Beryllium-ICP-Acid Dill.	ug/mL	102.8	<5.00e-03	<2.000	<2.00e0	n/a	n/a	n/a	2.000	n/a
S961003746	D	Bismuth-ICP-Acid Dill.	ug/mL	99.20	<1.00e-01	<40.10	<4.01e1	<40.10	102.0	n/a	40.10	n/a
S961003746	D	Calcium-ICP-Acid Dill.	ug/mL	101.2	<1.00e-01	<40.10	<4.01e1	<40.10	103.0	n/a	40.10	n/a
S961003746	D	Cadmium-ICP-Acid Dill.	ug/mL	100.0	<5.00e-03	<2.000	<2.00e0	n/a	n/a	n/a	2.000	n/a
S961003746	D	Cerium-ICP-Acid Dill.	ug/mL	99.00	<1.00e-01	<40.10	<4.01e1	<40.10	99.40	n/a	40.10	n/a
S961003746	D	Chromium-ICP-Acid Dill.	ug/mL	100.6	<2.00e-02	<8.020	<8.02e0	n/a	n/a	n/a	8.020	n/a
S961003746	D	Cobalt-ICP-Acid Dill.	ug/mL	100.6	<1.00e-02	2.12e+02	219.0	215.5	3.25	n/a	4.010	n/a
S961003746	D	Copper-ICP-Acid Dill.	ug/mL	103.4	<1.00e-02	<40.10	<4.01e1	<40.10	103.0	n/a	4.010	n/a
S961003746	D	Iron-ICP-Acid Dill.	ug/mL	101.8	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	20.10	n/a
S961003746	D	Potassium-ICP-Acid Dill.	ug/mL	99.60	<5.00e-02	6.19e+03	6.49e+03	6.34e+03	4.73	n/a	200.10	n/a
S961003746	D	Lanthanum-ICP-Acid Dill.	ug/mL	100.6	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	20.10	n/a
S961003746	D	Lithium-ICP-Acid Dill.	ug/mL	101.6	<1.00e-02	5.940	6.040	5.990	1.67	n/a	4.010	n/a
S961003746	D	Magnesium-ICP-Acid Dill.	ug/mL	97.20	<1.00e-01	<40.10	<4.01e1	n/a	n/a	n/a	40.10	n/a
S961003746	D	Manganese-ICP-Acid Dill.	ug/mL	97.60	<1.00e-02	<40.10	<4.01e1	n/a	n/a	n/a	4.010	n/a
S961003746	D	Molybdenum-ICP-Acid Dill.	ug/mL	100.2	<5.00e-02	94.50	96.90	95.70	2.51	n/a	20.10	n/a
S961003746	D	Sodium-ICP-Acid Dill.	ug/mL	98.00	<1.00e-01	2.40e+05	2.46e+05	2.43e+05	2.47	n/a	40.10	n/a
S961003746	D	Neodymium-ICP-Acid Dill.	ug/mL	101.2	<1.00e-01	<40.10	<4.01e1	n/a	n/a	n/a	40.10	n/a
S961003746	D	Nickel-ICP-Acid Dill.	ug/mL	106.4	<2.00e-02	<8.020	<8.02e0	n/a	n/a	n/a	8.020	n/a
S961003746	D	Phosphorus-ICP-Acid Dill.	ug/mL	99.40	<1.00e-01	4.17e+02	4.13e0	4.15e0	0.96	n/a	80.20	n/a
S961003746	D	Lead-ICP-Acid Dill.	ug/mL	99.40	<1.00e-01	52.40	49.50	50.95	5.69	n/a	40.10	n/a
S961003746	D	Sulfur-ICP-Acid Dill.	ug/mL	100.6	<1.00e-01	9.75e+02	986.0	980.5	1.12	n/a	40.10	n/a
S961003746	D	Antimony-ICP-Acid Dill.	ug/mL	95.60	<5.00e-02	<24.10	<2.41e1	n/a	n/a	n/a	24.10	n/a
S961003746	D	Selenium-ICP-Acid Dill.	ug/mL	97.60	<1.00e-01	<40.10	<4.01e1	n/a	n/a	n/a	40.10	n/a
S961003746	D	Silicon-ICP-Acid Dill.	ug/mL	98.80	<5.00e-02	2.02e+02	207.0	204.5	2.44	n/a	20.10	n/a
S961003746	D	Strontium-ICP-Acid Dill.	ug/mL	97.40	<1.00e-01	<40.10	<4.01e1	n/a	n/a	n/a	40.10	n/a
S961003746	D	Titanium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-02	<40.10	<4.01e1	n/a	n/a	n/a	4.010	n/a
S961003746	D	Thallium-ICP-Acid Dill.	ug/mL	97.20	<1.00e-02	<40.10	<4.01e1	n/a	n/a	n/a	4.010	n/a
S961003746	D	Uranium-ICP-Acid Dill.	ug/mL	96.00	<5.00e-01	<80.20	<8.02e1	n/a	n/a	n/a	80.20	n/a
S961003746	D	Vanadium-ICP-Acid Dill.	ug/mL	95.40	<5.00e-02	<20.10	<2.01e1	n/a	n/a	n/a	200.10	n/a
S961003746	D	Zinc-ICP-Acid Dill.	ug/mL	100.0	<1.00e-02	<33.20	<3.32e0	31.80	8.81	n/a	4.010	n/a
S961003746	D	Zirconium-ICP-Acid Dill.	ug/mL	98.00	<2.00e-02	4.050	<4.01e1	n/a	n/a	n/a	4.010	n/a
S961003746	D	Acetate by IC-Dipex 4000/4500	ug/mL	90.32	<2.00e-01	33.16	32.60	32.88	1.82	n/a	20.19	n/a
S961003746	D	Formate by Ion Chromatograph	ug/mL	98.33	<2.00e-01	52.30	51.60	51.95	1.35	n/a	20.19	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R /#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003746		Fluoride-IC-Dioxex 4000/4500	ug/mL	102.0	<1.30e-02	< 53.83	<5.38e1	n/a	n/a	106.6	53.83	n/a
S96T003746		Chloride-IC-Dioxex 4000/4500	ug/mL	102.7	<1.70e-02	9.61e+03	9.41e+03	9.51e+03	2.10	91.77	70.40	n/a
S96T003746		Nitrite-IC-Dioxex 4000/4500	ug/mL	97.24	<1.07e-01	1.03e+05	1.03e+05	1.03e+05	3.60	113.6	1.08e+03	n/a
S96T003746		Bromide by Ion Chromatography	ug/mL	98.47	<1.26e-01	7.18e+05	744.0	750.8	3.56	96.94	521.8	n/a
S96T003746		Nitrate by IC-Dioxex 4000/4500	ug/mL	99.83	2.21e-01	1.31e+05	1.30e+05	1.30e+05	0.77	113.5	1.41e+03	n/a
S96T003746		Phosphate-IC-Dioxex 4000/4500	ug/mL	106.2	<1.19e-01	<4.92e+02	<4.92e2	n/a	n/a	108.1	492.4	n/a
S96T003746		Sulfate by IC-Dioxex 4000/4500	ug/mL	99.84	<1.36e-01	1.16e+03	962.0	1.06e+03	18.7	96.99	562.8	n/a
S96T003746		Oxalate by IC-Dioxex 4000/4500	ug/mL	102.8	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	103.4	434.8	n/a
S96T003746		Alpha in Liquid Samples	uc/mL	100.0	<3.04e-02	<2.74e-02	<2.74E-2	n/a	n/a	n/a	3.70e-02	5.00E+02

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003747		DSC Exotherm using Mettler	Joules/g	94.20	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003747		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003747		Specific Gravity	Sp.G.	99.53	n/a	1.367	1.368	1.368	0.07	n/a	1.00e-03	n/a
S961003747		% Water by TGA on Perkin Elmer		99.53	n/a	50.12	49.66	49.89	0.92	n/a	n/a	n/a
S961003747		D Silicon-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	17.20	16.90	17.05	1.76	93.00	6.010	n/a
S961003747		D Aluminum-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	4.16e+04	4.24e+04	4.20e+04	1.90	70.20	30.10	n/a
S961003747		D Boron-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	94.60	60.10	n/a
S961003747		D Barium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 30.10	<3.01e1	< 30.10	3.82	96.00	30.10	n/a
S961003747		D Beryllium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	< 3.000	<3.00e0	< 3.000	n/a	97.60	30.10	n/a
S961003747		D Bismuth-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	99.70	3.000	n/a
S961003747		D Cadmium-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	99.00	60.10	n/a
S961003747		D Cerium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-03	< 3.000	<3.00e0	< 3.000	n/a	101.0	3.000	n/a
S961003747		D Cobalt-ICP-Acid Dil.	ug/mL	101.6	<2.00e-02	< 12.00	<1.20e1	< 12.00	n/a	99.50	60.10	n/a
S961003747		D Chromium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	2.13e+02	218.0	215.5	2.32	101.0	6.010	n/a
S961003747		D Copper-ICP-Acid Dil.	ug/mL	103.2	<1.00e-02	< 6.010	<6.01e0	< 6.010	n/a	101.0	6.010	n/a
S961003747		D Iron-ICP-Acid Dil.	ug/mL	103.0	<5.00e-02	< 30.10	<3.01e1	< 30.10	n/a	101.0	30.10	n/a
S961003747		D Potassium-ICP-Acid Dil.	ug/mL	94.80	<5.00e-01	6.49e+03	6.67e+03	6.58e+03	2.74	106.0	300.0	n/a
S961003747		D Lanthanum-ICP-Acid Dil.	ug/mL	102.0	<5.00e-02	< 30.10	<3.01e1	< 30.10	n/a	101.0	30.10	n/a
S961003747		D Lithium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	< 6.010	<6.01e0	< 6.010	n/a	103.0	6.010	n/a
S961003747		D Magnesium-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	< 60.10	<6.01e0	< 60.10	n/a	94.80	60.10	n/a
S961003747		D Manganese-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	< 6.010	<6.01e0	< 6.010	n/a	96.50	6.010	n/a
S961003747		D Molybdenum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	98.80	98.20	98.50	0.61	101.0	30.10	n/a
S961003747		D Sodium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	2.44e+05	2.48e+05	2.46e+05	1.63	-1.17e+02	60.10	n/a
S961003747		D Neodymium-ICP-Acid Dil.	ug/mL	102.6	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	102.0	60.10	n/a
S961003747		D Nickel-ICP-Acid Dil.	ug/mL	101.6	<2.00e-02	< 12.00	<1.20e1	< 12.00	n/a	101.0	12.00	n/a
S961003747		D Phosphorus-ICP-Acid Dil.	ug/mL	99.00	<2.00e-01	3.27e+02	331.0	329.0	1.22	101.0	120.0	n/a
S961003747		D Lead-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	< 60.10	<6.01e0	< 60.10	n/a	112.0	60.10	n/a
S961003747		D Sulphur-ICP-Acid Dil.	ug/mL	98.60	<1.00e-01	9.49e+02	973.0	961.0	2.50	101.0	60.10	n/a
S961003747		D Antimony-ICP-Acid Dil.	ug/mL	95.40	<1.00e-02	< 36.10	<3.61e1	< 36.10	n/a	99.50	36.10	n/a
S961003747		D Selenium-ICP-Acid Dil.	ug/mL	96.20	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	97.80	60.10	n/a
S961003747		D Silicon-ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	1.93e+02	197.0	195.0	2.05	101.0	30.10	n/a
S961003747		D Strontium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-01	< 60.10	<6.01e1	< 60.10	n/a	97.70	60.10	n/a
S961003747		D Titanium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-02	< 6.010	<6.01e0	< 6.010	n/a	96.60	6.010	n/a
S961003747		D Thallium-ICP-Acid Dil.	ug/mL	97.00	<2.00e-01	<3.00e+02	<3.00e2	<3.00e+02	n/a	93.90	300.0	n/a
S961003747		D Uranium-ICP-Acid Dil.	ug/mL	96.50	<5.00e-01	<3.00e+02	<3.01e1	<3.00e+02	n/a	99.00	30.10	n/a
S961003747		D Vanadium-ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	< 30.10	<3.01e1	< 30.10	n/a	103.0	6.010	n/a
S961003747		D Zinc-ICP-Acid Dil.	ug/mL	102.4	<1.00e-02	< 6.010	<6.01e0	< 6.010	6.41	99.70	6.010	n/a
S961003747		D Zirconium-ICP-Acid Dil.	ug/mL	99.32	<2.00e-01	5.32e+02	538.0	534.9	1.12	n/a	222.1	n/a
S961003747		Acetate by IC-Dioxen 4000/4500	ug/mL	98.33	<2.00e-01	1.36e+03	1.37e+03	1.37e+03	0.73	n/a	222.1	n/a
S961003747		Formate by Ion Chromatograph	ug/mL	98.33	<2.00e-01	1.36e+03	1.37e+03	1.37e+03	0.73	n/a	222.1	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003747		Fluoride-IC-Dionex 4000/4500	ug/mL	102.0	<1.50e-02	9.24e-02	<1.38e-1	n/a	n/a	n/a	53.83	n/a
S961003747		Chloride-IC-Dionex 4000/4500	ug/mL	102.7	<1.70e-02	9.68e-03	9.51e-03	9.58e-03	1.56	n/a	70.40	n/a
S961003747		Nitrite-IC-Dionex 4000/4500	ug/mL	97.24	<1.07e-01	1.23e-03	1.23e-03	1.23e-03	0.100	n/a	1.08e+03	n/a
S961003747		Bromide by Ion Chromatograph	ug/mL	104.6	<1.28e-01	1.08e-03	1.04e-03	1.06e-03	3.77	99.15	649.0	n/a
S961003747		Nitrate by IC-Dionex 4000/4500	ug/mL	99.83	2.21e-01	1.58e-05	1.36e-05	1.57e-05	1.27	n/a	1.41e+03	n/a
S961003747		Phosphate-IC-Dionex 4000/4500	ug/mL	106.2	<1.19e-01	1.73e-04	<4.92e-2	n/a	n/a	n/a	492.4	n/a
S961003747		Sulfate by IC-Dionex 4000/4500	ug/mL	102.9	<1.36e-01	1.14e-03	1.11e-03	1.12e-03	2.67	101.7	700.0	n/a
S961003747		Oxalate by IC-Dionex 4000/4500	ug/mL	102.8	<1.05e-01	1.61e-04	<4.35e-2	n/a	n/a	n/a	434.8	n/a
S961003747		Alpha in Liquid Samples	ug/L/mL	100.0	<3.04e-02	<3.33e-02	<1.52e-2	n/a	n/a	n/a	3.70e-02	5.00E+02

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Table 1. Data Summary Table
AN-105

CORE NUMBER: 152
SEGMENT #: 11

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec.	Det Limit	Count Err%
S96T003748		DSC Exotherm using Mettler	Joules/g	100.9	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003748		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003748		Specific Gravity	Sp.G.	98.74	n/a	1.382	1.382	1.382	0.00	n/a	1.00e-03	n/a
S96T003748		% Water by TGA using Mettler	%	98.90	n/a	49.75	30.12	39.94	49.2	n/a	n/a	n/a
S96T003748	1	% Water by TGA using Mettler	%	98.90	n/a	50.51	50.66	50.58	0.30	n/a	n/a	n/a
S96T003748		Aluminum-ICP-Acid Dill.	ug/mL	100.6	<1.00e-02	17.20	17.70	17.45	2.87	92.70	6.010	n/a
S96T003748		Aluminum-ICP-Acid Dill.	ug/mL	98.20	<1.00e-02	4.15e+04	4.20e+04	4.18e+04	1.20	118.0	30.10	n/a
S96T003748		Aluminum-ICP-Acid Dill.	ug/mL	102.8	<1.00e-02	60.10	<6.01e1	n/a	n/a	94.80	60.10	n/a
S96T003748		Boron-ICP-Acid Dill.	ug/mL	100.0	<5.00e-02	76.20	78.90	77.55	3.48	97.20	30.10	n/a
S96T003748		Boron-ICP-Acid Dill.	ug/mL	100.2	<5.00e-02	30.10	<3.01e1	n/a	n/a	97.80	30.10	n/a
S96T003748		Beryllium-ICP-Acid Dill.	ug/mL	101.6	<5.00e-03	3.000	<3.00e0	n/a	n/a	99.90	3.000	n/a
S96T003748		Bismuth-ICP-Acid Dill.	ug/mL	102.8	<1.00e-01	60.10	<6.01e1	n/a	n/a	101.0	60.10	n/a
S96T003748		Calcium-ICP-Acid Dill.	ug/mL	102.0	<1.00e-01	60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a
S96T003748		Cadmium-ICP-Acid Dill.	ug/mL	100.4	<5.00e-03	3.000	<3.00e0	n/a	n/a	101.0	3.000	n/a
S96T003748		Cerium-ICP-Acid Dill.	ug/mL	101.6	<1.00e-01	60.10	<6.01e1	n/a	n/a	100.0	60.10	n/a
S96T003748		Cobalt-ICP-Acid Dill.	ug/mL	102.0	<1.00e-02	12.00	<1.20e1	n/a	n/a	101.0	12.00	n/a
S96T003748		Chromium-ICP-Acid Dill.	ug/mL	102.0	<1.00e-02	2.12e+02	214.0	213.0	0.94	101.0	6.010	n/a
S96T003748		Copper-ICP-Acid Dill.	ug/mL	103.2	<1.00e-02	6.010	<6.01e0	n/a	n/a	101.0	6.010	n/a
S96T003748		Iron-ICP-Acid Dill.	ug/mL	103.0	<5.00e-02	30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a
S96T003748		Potassium-ICP-Acid Dill.	ug/mL	94.80	<5.00e-01	6.49e+03	6.63e+03	6.56e+03	2.13	138.0	300.0	n/a
S96T003748		Lanthanum-ICP-Acid Dill.	ug/mL	102.0	<5.00e-02	30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a
S96T003748		Lithium-ICP-Acid Dill.	ug/mL	101.6	<1.00e-02	6.010	<6.01e0	n/a	n/a	103.0	6.010	n/a
S96T003748		Magnesium-ICP-Acid Dill.	ug/mL	99.00	<1.00e-01	60.10	<6.01e1	n/a	n/a	93.30	60.10	n/a
S96T003748		Manganese-ICP-Acid Dill.	ug/mL	96.00	<1.00e-02	6.010	<6.01e0	n/a	n/a	96.20	6.010	n/a
S96T003748		Molybdenum-ICP-Acid Dill.	ug/mL	101.6	<5.00e-02	94.90	98.40	96.65	3.62	102.0	30.10	n/a
S96T003748		Sodium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-01	2.42e+05	2.44e+05	2.43e+05	0.82	503.0	60.10	n/a
S96T003748		Neodymium-ICP-Acid Dill.	ug/mL	102.6	<1.00e-01	60.10	<6.01e1	n/a	n/a	101.0	60.10	n/a
S96T003748		Nickel-ICP-Acid Dill.	ug/mL	101.6	<1.00e-02	12.00	<1.20e1	n/a	n/a	101.0	12.00	n/a
S96T003748		Phosphorus-ICP-Acid Dill.	ug/mL	99.00	<2.00e-02	3.29e+02	311.0	320.0	5.62	95.20	120.0	n/a
S96T003748		Lead-ICP-Acid Dill.	ug/mL	101.4	<1.00e-01	60.10	<6.01e1	n/a	n/a	110.0	60.10	n/a
S96T003748		Sulfur-ICP-Acid Dill.	ug/mL	98.60	<1.00e-01	9.62e+02	975.0	968.5	1.34	94.40	60.10	n/a
S96T003748		Antimony-ICP-Acid Dill.	ug/mL	99.40	<5.00e-02	36.10	<3.61e1	n/a	n/a	97.40	36.10	n/a
S96T003748		Selenium-ICP-Acid Dill.	ug/mL	96.20	<1.00e-01	60.10	<6.01e1	n/a	n/a	97.90	60.10	n/a
S96T003748		Silicon-ICP-Acid Dill.	ug/mL	99.60	<1.00e-02	2.30e+02	231.0	230.5	0.43	101.0	30.10	n/a
S96T003748		Samarium-ICP-Acid Dill.	ug/mL	98.20	<1.00e-01	60.10	<6.01e1	n/a	n/a	97.80	60.10	n/a
S96T003748		Strontium-ICP-Acid Dill.	ug/mL	99.40	<1.00e-02	6.010	<6.01e0	n/a	n/a	97.70	6.010	n/a
S96T003748		Titanium-ICP-Acid Dill.	ug/mL	97.80	<1.00e-02	6.010	<6.01e0	n/a	n/a	95.30	120.0	n/a
S96T003748		Thallium-ICP-Acid Dill.	ug/mL	97.00	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	98.00	300.0	n/a
S96T003748		Uranium-ICP-Acid Dill.	ug/mL	96.50	<5.00e-02	<3.00e+02	<3.00e2	n/a	n/a	100.0	30.10	n/a
S96T003748		Vanadium-ICP-Acid Dill.	ug/mL	101.0	<1.00e-02	30.10	<3.01e1	n/a	n/a	102.0	6.010	n/a
S96T003748		Zinc-ICP-Acid Dill.	ug/mL	102.4	<1.00e-02	16.10	16.20	16.15	0.62	102.0	6.010	n/a
S96T003748		Zirconium-ICP-Acid Dill.	ug/mL	99.60	<1.00e-02	6.090	<6.01e0	n/a	n/a	98.70	6.010	n/a
S96T003748		Acetate by IC-Dioxen 4000/4500	ug/mL	108.1	<2.00e-02	4.40e+02	407.0	423.4	7.79	n/a	204.0	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961003748		Formate by Ion Chromatograph	ug/mL	95.00	<2.10e-02	1.50e+03	1.53e+03	1.51e+03	1.98	n/a	214.2		n/a
S961003748		Fluoride-IC-Dionex 4000/4500	ug/mL	101.7	<1.30e-02	4.47e+02	468.0	457.5	4.59	n/a	27.57		n/a
S961003748		Chloride-IC-Dionex 4000/4500	ug/mL	101.4	3.20e-02	1.01e+04	1.01e+04	1.01e+04	0.00	n/a	36.06		n/a
S961003748		Nitrite-IC - Dionex 4000/4500	ug/mL	96.81	<1.07e-01	1.26e+05	1.25e+05	1.25e+05	0.80	n/a	551.2		n/a
S961003748		Bromide by Ion Chromatograph	ug/mL	100.2	<1.26e-01	6.39e+02	654.0	646.7	2.32	n/a	267.2		n/a
S961003748		Nitrate by IC-Dionex 4000/4500	ug/mL	100.3	1.54e-01	1.64e+05	1.64e+05	1.64e+05	0.00	n/a	721.1		n/a
S961003748		Phosphate-IC-Dionex 4000/4500	ug/mL	99.45	<1.19e-01	7.09e+02	615.0	662.0	14.2	n/a	252.2		n/a
S961003748		Sulfate by IC-Dionex 4000/4500	ug/mL	100.6	<1.36e-01	8.56e+02	944.0	900.1	9.78	n/a	288.2		n/a
S961003748		Oxalate by IC-Dionex 4000/4500	ug/mL	95.07	<1.03e-01	2.57e+02	270.0	263.7	4.93	n/a	222.7		n/a
S961003748		Alpha in Liquid Samples	uc1/mL	93.75	<5.59e-02	<5.59e-02	<8.27E-2	n/a	n/a	122.3	9.70e-02	5.00E+02	

CORE NUMBER: 152
SEGMENT #: 12

SEGMENT PORTION: Drainable Liquid

Table 1. Data Summary Table
AN-105

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003749 1		DSC Exotherm DRY, Calculated	Joules/g DRY	n/a	n/a	85.70	131.0	108.3	41.8	n/a	n/a	n/a
S96T003749		DSC Exotherm on Perkin Elmer	Joules/g	96.34	n/a	42.60	64.90	53.75	41.5	n/a	n/a	n/a
S96T003749		Specific Gravity	Sp.G.	98.74	n/a	1.392	1.383	1.387	0.75	n/a	1.00e-03	n/a
S96T003749		% Water by IGA using Mettler	%	98.90	n/a	50.07	50.46	50.27	0.68	n/a	n/a	n/a
S96T003749		Silver-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	17.90	16.80	17.35	6.34	92.20	6.010	n/a
S96T003749		Aluminum-ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	4.14e+04	4.03e+04	4.08e+04	2.69	-8.35e+01	30.10	n/a
S96T003749		Arsenic-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	94.40	60.10	n/a
S96T003749		Boron-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	77.10	75.00	76.05	2.76	97.60	30.10	n/a
S96T003749		Barium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	98.20	30.10	n/a
S96T003749		Beryllium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	100.0	< 3.000	n/a
S96T003749		Bismuth-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	99.60	60.10	n/a
S96T003749		Calcium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	103.0	< 60.10	n/a
S96T003749		Cadmium-ICP-Acid Dil.	ug/mL	101.4	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	101.0	< 3.000	n/a
S96T003749		Cerium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	100.0	< 60.10	n/a
S96T003749		Cobalt-ICP-Acid Dil.	ug/mL	101.6	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	100.0	< 12.00	n/a
S96T003749		Chromium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	2.14e+02	208.0	211.0	2.84	100.0	< 6.010	n/a
S96T003749		Copper-ICP-Acid Dil.	ug/mL	103.2	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	101.0	< 6.010	n/a
S96T003749		Iron-ICP-Acid Dil.	ug/mL	103.0	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	100.0	< 30.10	n/a
S96T003749		Potassium-ICP-Acid Dil.	ug/mL	94.80	<5.00e-01	6.60e+03	6.42e+03	6.51e+03	2.76	79.00	300.0	n/a
S96T003749		Lanthanum-ICP-Acid Dil.	ug/mL	102.0	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	101.0	< 30.10	n/a
S96T003749		Lithium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	6.700	6.620	6.660	1.20	102.0	< 6.010	n/a
S96T003749		Magnesium-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	94.00	< 60.10	n/a
S96T003749		Manganese-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	96.40	< 6.010	n/a
S96T003749		Molybdenum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	98.10	94.50	96.30	3.74	101.0	< 30.10	n/a
S96T003749		Sodium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	2.40e+01	2.34e+01	2.37e+01	2.53	-7.99e+02	60.10	n/a
S96T003749		Neodymium-ICP-Acid Dil.	ug/mL	102.6	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	102.0	< 60.10	n/a
S96T003749		Nickel-ICP-Acid Dil.	ug/mL	99.00	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	100.0	< 12.00	n/a
S96T003749		Phosphorus-ICP-Acid Dil.	ug/mL	101.6	<2.00e-02	3.28e+02	323.0	325.5	1.54	98.40	< 120.0	n/a
S96T003749		Lead-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	111.0	< 60.10	n/a
S96T003749		Sulfur-ICP-Acid Dil.	ug/mL	98.60	<1.00e-01	9.55e+02	923.0	939.0	3.41	94.70	< 60.10	n/a
S96T003749		Antimony-ICP-Acid Dil.	ug/mL	96.40	<1.00e-02	< 36.10	<3.61e1	n/a	n/a	96.80	< 36.10	n/a
S96T003749		Selenium-ICP-Acid Dil.	ug/mL	96.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	100.0	< 60.10	n/a
S96T003749		Silicon-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	2.07e+02	201.0	204.0	2.94	99.50	< 30.10	n/a
S96T003749		Samarium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	97.60	< 60.10	n/a
S96T003749		Strontium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	98.20	< 6.010	n/a
S96T003749		Titanium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	96.50	< 6.010	n/a
S96T003749		Thallium-ICP-Acid Dil.	ug/mL	97.00	<5.00e-01	<1.20e+02	<1.20e+02	<3.00e2	n/a	93.40	< 120.0	n/a
S96T003749		Uranium-ICP-Acid Dil.	ug/mL	96.50	<2.00e-01	<3.00e+02	<3.00e2	n/a	n/a	98.00	< 300.0	n/a
S96T003749		Vanadium-ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	100.0	< 30.10	n/a
S96T003749		Zinc-ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	16.30	16.00	16.15	1.86	102.0	< 6.010	n/a
S96T003749		Dicromate by IC-Dioxex	ug/mL	99.60	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	99.90	< 6.010	n/a
S96T003749		Acetate by IC-Dioxex	ug/mL	108.1	<2.00e-02	4.13e+02	453.0	433.2	9.24	121.8	204.0	n/a
S96T003749		Formate by Ion Chromatograph	ug/mL	95.00	<2.10e-02	1.70e+03	1.75e+03	1.73e+03	2.90	94.17	214.2	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003749		Fluoride-IC-Dionex 4000/4500	µg/mL	101.7	<1.30e-02	4.68e+02	462.0	465.1	1.29	85.76	66.96	n/a
S961003749		Chloride-IC-Dionex 4000/4500	µg/mL	101.4	3.20e-02	9.83e+03	9.89e+03	9.86e+03	0.61	100.6	87.57	n/a
S961003749		Nitrite-IC - Dionex 4000/4500	µg/mL	96.81	<1.07e-01	1.81e+03	1.25e+03	1.23e+03	3.25	105.1	551.2	n/a
S961003749		Bromide by Ion Chromatograph	µg/mL	100.2	<1.20e-01	1.02e+03	1.04e+03	1.03e+03	1.94	92.87	649.0	n/a
S961003749		Nitrate by IC-Dionex 4000/4500	µg/mL	100.3	1.54e-01	1.32e+03	1.35e+03	1.34e+03	1.95	104.0	721.1	n/a
S961003749		Phosphate-IC-Dionex 4000/4500	µg/mL	99.45	<1.39e-01	6.87e+02	<6.12e2	n/a	n/a	100.4	612.5	n/a
S961003749		Sulfate by IC-Dionex 4000/4500	µg/mL	100.6	<1.38e-01	1.07e+03	1.14e+03	1.10e+03	6.33	100.5	700.0	n/a
S961003749		Oxalate by IC-Dionex 4000/4500	µg/mL	95.07	<1.05e-01	<5.41e+02	<5.41e2	n/a	n/a	96.20	540.9	n/a
S961003749		Alpha in Liquid Samples	UC/mL	93.75	<5.59e-02	<5.59e-02	<5.59E-2	n/a	n/a	n/a	9.70e-02	5.00E+02

HNF-SD-WV-DP-199, REV. 1

W-IC-SD-WM-DP-199, REV 1

CORE NUMBER: 152
SEGMENT #: 13

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003750		DSC Exotherm using Mettler	Joules/g	102.6	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003750		DSC Exotherm DRY Calculated	Joules/g DRY		n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003750		% Specific Gravity	Sp.G.	99.31	n/a	1.376	1.388	1.382	0.87	n/a	1.00e-03	n/a
S961003750		% Water by TGA on perkin Elmer	%	99.31	n/a	48.76	50.22	49.49	2.95	n/a	n/a	n/a
S961003750		D Silver-ICP-Acid Dill.	ug/mL	100.6	<1.00e-02	16.70	16.80	16.75	0.60	91.40	6.010	n/a
S961003750		D Aluminum-ICP-Acid Dill.	ug/mL	98.20	<1.00e-02	4.01e+04	3.98e+04	4.00e+04	0.75	187.0	30.10	n/a
S961003750		D Arsenic-ICP-Acid Dill.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	71.70	n/a	94.90	60.10	n/a
S961003750		D Boron-ICP-Acid Dill.	ug/mL	100.0	<1.00e-02	68.80	67.61	70.25	4.13	98.50	30.10	n/a
S961003750		D Barium-ICP-Acid Dill.	ug/mL	100.2	<1.00e-02	< 3.000	<3.01e1	n/a	n/a	98.30	30.10	n/a
S961003750		D Beryllium-ICP-Acid Dill.	ug/mL	101.6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	100.0	3.000	n/a
S961003750		D Bismuth-ICP-Acid Dill.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	94.80	60.10	n/a
S961003750		D Calcium-ICP-Acid Dill.	ug/mL	102.0	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a
S961003750		D Cadmium-ICP-Acid Dill.	ug/mL	101.4	<5.00e-03	< 3.000	<3.00e0	101.4	n/a	101.0	3.000	n/a
S961003750		D Cerium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	100.0	60.10	n/a
S961003750		D Cobalt-ICP-Acid Dill.	ug/mL	101.6	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	102.0	12.00	n/a
S961003750		D Chromium-ICP-Acid Dill.	ug/mL	102.0	<1.00e-02	2.06e+02	202.0	204.0	1.96	101.0	6.010	n/a
S961003750		D Copper-ICP-Acid Dill.	ug/mL	103.2	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	101.0	6.010	n/a
S961003750		D Iron-ICP-Acid Dill.	ug/mL	103.0	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	100.0	30.10	n/a
S961003750		D Potassium-ICP-Acid Dill.	ug/mL	94.80	<5.00e-02	6.41e+03	6.31e+03	6.36e+03	1.57	130.0	300.0	n/a
S961003750		D Lanthanum-ICP-Acid Dill.	ug/mL	102.0	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a
S961003750		D Lithium-ICP-Acid Dill.	ug/mL	101.6	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	103.0	6.010	n/a
S961003750		D Magnesium-ICP-Acid Dill.	ug/mL	96.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	94.90	60.10	n/a
S961003750		D Manganese-ICP-Acid Dill.	ug/mL	99.00	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	96.70	6.010	n/a
S961003750		D Molybdenum-ICP-Acid Dill.	ug/mL	101.6	<5.00e-02	93.40	91.10	92.25	2.49	102.0	30.10	n/a
S961003750		D Sodium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-01	2.38e+05	2.36e+05	2.37e+05	0.84	671.0	60.10	n/a
S961003750		D Neodymium-ICP-Acid Dill.	ug/mL	102.6	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a
S961003750		D Nickel-ICP-Acid Dill.	ug/mL	101.6	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	101.0	12.00	n/a
S961003750		D Phosphorus-ICP-Acid Dill.	ug/mL	99.00	<2.00e-01	3.21e+02	316.0	318.5	1.57	99.10	120.0	n/a
S961003750		D Lead-ICP-Acid Dill.	ug/mL	101.4	<1.00e-01	61.10	<6.01e1	n/a	n/a	103.0	60.10	n/a
S961003750		D Sulfur-ICP-Acid Dill.	ug/mL	98.60	<1.00e-01	9.33e+02	928.0	930.5	0.54	101.0	60.10	n/a
S961003750		D Antimony-ICP-Acid Dill.	ug/mL	95.40	<1.00e-02	< 36.10	<3.61e1	n/a	n/a	97.10	36.10	n/a
S961003750		D Selenium-ICP-Acid Dill.	ug/mL	96.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a
S961003750		D Silicon-ICP-Acid Dill.	ug/mL	99.60	<5.00e-02	2.04e+02	205.0	204.5	0.49	102.0	30.10	n/a
S961003750		D Samarium-ICP-Acid Dill.	ug/mL	98.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	97.40	60.10	n/a
S961003750		D Strontium-ICP-Acid Dill.	ug/mL	99.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	98.10	6.010	n/a
S961003750		D Thallium-ICP-Acid Dill.	ug/mL	97.80	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	96.80	6.010	n/a
S961003750		D Uranium-ICP-Acid Dill.	ug/mL	97.00	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	94.50	120.0	n/a
S961003750		D Vanadium-ICP-Acid Dill.	ug/mL	96.50	<5.00e-01	< 30.10	<3.01e1	n/a	n/a	98.00	300.0	n/a
S961003750		D Zirconium-ICP-Acid Dill.	ug/mL	101.0	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	102.0	6.010	n/a
S961003750		D Zinc-ICP-Acid Dill.	ug/mL	102.4	<1.00e-02	14.70	14.70	14.70	0.00	100.0	30.10	n/a
S961003750		D Chromium-ICP-Acid Dill.	ug/mL	99.60	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	99.90	6.010	n/a
S961003750		D Acetate by IC-Dipnax 4000/4500	ug/mL	112.1	<2.00e-02	6.02e+02	599.0	600.5	0.50	n/a	103.0	n/a
S961003750		D Formate by Ion Chromatograph	ug/mL	99.17	<2.10e-02	1.61e+03	1.61e+03	1.61e+03	0.00	n/a	108.2	n/a

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S9610038704		Bulk Density of Sample	g/mL	n/a	n/a	1.530	n/a	n/a	n/a	5.00e-01	n/a
S9610038116		DSC Exotherm using Mettler	Joules/g	98.77	n/a	28.40	24.20	26.30	16.0	n/a	n/a
S9610038116		DSC Exotherm DRY Calculated	Joules/g Dry	n/a	n/a	53.50	45.80	49.55	15.9	n/a	n/a
S9610038116		% Water by IGA using Mettler	%	99.51	n/a	47.82	46.02	46.92	3.84	n/a	n/a
S961003894		F Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.00e-02	<2.00e-02	2.56e+04	2.36	94.10	204.0
S961003894		F Aluminum - ICP-Fusion	ug/g	98.20	<5.00e-02	<2.50e-02	<2.50e-02	2.56e+04	2.36	94.10	1.02e+03
S961003894		F Arsenic - ICP-Fusion	ug/g	103.4	<1.00e-01	<2.00e-03	<2.00e-03	n/a	n/a	101.0	1.02e+03
S961003894		F Boron - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	93.70	1.02e+03
S961003894		F Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	94.00	1.02e+03
S961003894		F Beryllium - ICP-Fusion	ug/g	103.6	<5.00e-03	<1.02e+02	<1.02e+02	n/a	n/a	98.10	102.0
S961003894		F Bismuth - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.00e+03	<2.00e+03	n/a	n/a	99.00	2.04e+03
S961003894		F Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<1.02e+03	<1.02e+03	n/a	n/a	98.50	2.04e+03
S961003894		F Cadmium - ICP-Fusion	ug/g	100.0	<5.00e-03	<1.02e+02	<1.02e+02	n/a	n/a	97.90	102.0
S961003894		F Cerium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.00e+03	<2.00e+03	n/a	n/a	95.60	2.04e+03
S961003894		F Cobalt - ICP-Fusion	ug/g	100.8	<2.00e-02	<4.07e+02	<3.99e+02	n/a	n/a	97.20	408.0
S961003894		F Chromium - ICP-Fusion	ug/g	101.4	<1.00e-02	<1.23e+03	<1.23e+03	1.23e+03	0.00	98.50	204.0
S961003894		F Copper - ICP-Fusion	ug/g	103.4	<1.00e-02	<1.02e+03	<2.00e+02	n/a	n/a	97.20	204.0
S961003894		F Iron - ICP-Fusion	ug/g	101.0	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	97.50	1.02e+03
S961003894		F Potassium - ICP-Fusion	ug/g	96.80	<5.00e-01	<1.02e+03	<1.02e+03	n/a	n/a	97.50	1.02e+03
S961003894		F Lanthanum - ICP-Fusion	ug/g	101.6	<5.00e-02	<2.04e+02	<2.00e+02	n/a	n/a	95.00	204.0
S961003894		F Lithium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	93.70	2.04e+03
S961003894		F Magnesium - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.04e+02	<2.00e+02	n/a	n/a	93.00	204.0
S961003894		F Manganese - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	97.50	1.02e+03
S961003894		F Molybdenum - ICP-Fusion	ug/g	98.80	<1.00e-01	<2.04e+03	<1.02e+03	2.27e+05	9.69	98.00	2.04e+03
S961003894		F Sodium - ICP-Fusion	ug/g	101.8	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	96.30	2.04e+03
S961003894		F Neodymium - ICP-Fusion	ug/g	100.4	<1.24e-01	<6.19e+03	<1.71e+03	3.99e+03	113	96.30	408.0
S961003894		F Nickel - ICP-Fusion	ug/g	101.6	<2.00e-01	<4.07e+03	<3.99e+03	n/a	n/a	102.0	4.08e+03
S961003894		F Phosphorus - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	96.00	2.04e+03
S961003894		F Lead - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	107.0	2.04e+03
S961003894		F Sulfur - ICP-Fusion	ug/g	95.60	<6.00e-02	<1.22e+03	<1.20e+03	n/a	n/a	92.50	1.22e+03
S961003894		F Antimony - ICP-Fusion	ug/g	97.80	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	96.90	2.04e+03
S961003894		F Selenium - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	95.10	1.02e+03
S961003894		F Silicon - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.04e+03	<2.00e+03	n/a	n/a	94.70	2.04e+03
S961003894		F Samarium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.04e+02	<2.00e+02	n/a	n/a	94.50	204.0
S961003894		F Strontium - ICP-Fusion	ug/g	97.40	<1.00e-02	<2.04e+02	<2.00e+02	n/a	n/a	93.60	204.0
S961003894		F Titanium - ICP-Fusion	ug/g	96.80	<2.00e-01	<4.07e+03	<3.99e+03	n/a	n/a	93.30	4.08e+03
S961003894		F Thallium - ICP-Fusion	ug/g	96.00	<5.00e-01	<1.02e+04	<1.02e+04	n/a	n/a	96.70	1.02e+03
S961003894		F Uranium - ICP-Fusion	ug/g	100.2	<5.00e-02	<1.02e+03	<1.02e+03	n/a	n/a	97.00	1.02e+03
S961003894		F Vanadium - ICP-Fusion	ug/g	101.2	<1.00e-02	<3.97e+02	<3.97e+02	506.5	43.2	98.10	204.0
S961003894		F Zinc - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.04e+02	<2.00e+02	n/a	n/a	94.60	204.0
S961003894		F Zirconium - ICP-Fusion	ug/g	23.40	<1.00e-02	<13.80	<13.80	13.70	1.46	n/a	3.940
S961003898		A Silver - ICP-Acid Digest	ug/g	101.4	<2.28e-01	<2.51e+04	<2.59e+04	2.59e+04	3.14	n/a	19.70
S961003898		A Aluminum - ICP-Acid Digest	ug/g	101.4	<2.28e-01	<2.51e+04	<2.59e+04	2.59e+04	3.14	n/a	19.70

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003898	A	Arsenic - ICP-Acid Digest	ug/g	95.00	< 1.00e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Boron - ICP-Acid Digest	ug/g	118.2	6.30e-01	84.60	92.70	88.65	9.14	n/a	19.70	n/a
S961003898	A	Barium - ICP-Acid Digest	ug/g	92.80	< 5.00e-02	< 19.70	< 1.94e1	n/a	n/a	n/a	19.70	n/a
S961003898	A	Beryllium - ICP-Acid Digest	ug/g	94.60	< 5.00e-03	< 1.970	< 1.94e0	n/a	n/a	n/a	1.970	n/a
S961003898	A	Bismuth - ICP-Acid Digest	ug/g	92.40	1.30e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	2.14e+02	158.0	186.0	30.1	n/a	39.40	n/a
S961003898	A	Cadmium - ICP-Acid Digest	ug/g	90.80	< 5.00e-03	3.730	5.390	5.560	6.12	n/a	1.970	n/a
S961003898	A	Cerium - ICP-Acid Digest	ug/g	95.00	< 1.00e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Cobalt - ICP-Acid Digest	ug/g	93.80	< 2.00e-02	< 7.860	< 7.76e0	n/a	n/a	n/a	7.860	n/a
S961003898	A	Chromium - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	1.23e+03	1.25e+03	1.25e+03	3.20	n/a	3.940	n/a
S961003898	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	7.290	5.880	6.38	5.44	n/a	19.70	n/a
S961003898	A	Potassium - ICP-Acid Digest	ug/g	93.80	< 5.00e-01	3.90e+03	3.96e+03	3.94e+03	2.03	n/a	19.70	n/a
S961003898	A	Lanthanum - ICP-Acid Digest	ug/g	95.00	< 5.00e-02	< 19.70	< 1.94e1	n/a	n/a	n/a	19.70	n/a
S961003898	A	Lithium - ICP-Acid Digest	ug/g	88.60	< 1.00e-02	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Magnesium - ICP-Acid Digest	ug/g	90.00	< 1.00e-02	4.060	4.140	4.100	1.95	n/a	3.940	n/a
S961003898	A	Manganese - ICP-Acid Digest	ug/g	94.60	< 5.00e-02	58.90	60.00	59.45	1.85	n/a	39.40	n/a
S961003898	A	Sodium - ICP-Acid Digest	ug/g	150.6	1.900	1.79e+05	1.81e+05	1.80e+05	1.11	n/a	39.40	n/a
S961003898	A	Neodymium - ICP-Acid Digest	ug/g	94.00	< 1.00e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Nickel - ICP-Acid Digest	ug/g	93.40	< 2.00e-02	20.50	20.30	20.40	0.98	n/a	7.860	n/a
S961003898	A	Phosphorus - ICP-Acid Digest	ug/g	93.40	< 2.00e-01	3.58e+02	297.0	327.5	18.6	n/a	39.40	n/a
S961003898	A	Lead - ICP-Acid Digest	ug/g	89.00	< 1.00e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Sulfur - ICP-Acid Digest	ug/g	93.00	1.19e-01	2.03e+03	2.08e+03	2.06e+03	2.43	n/a	39.40	n/a
S961003898	A	Antimony - ICP-Acid Digest	ug/g	88.40	< 6.00e-02	< 23.60	< 2.33e1	n/a	n/a	n/a	23.60	n/a
S961003898	A	Selenium - ICP-Acid Digest	ug/g	87.40	< 1.00e-01	< 39.30	< 3.88e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Silicon - ICP-Acid Digest	ug/g	244.0	5.31e-01	1.38e+02	145.0	141.5	4.95	n/a	19.70	n/a
S961003898	A	Samarium - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	< 3.930	< 3.88e0	n/a	n/a	n/a	39.40	n/a
S961003898	A	Strontium - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	< 3.930	< 3.88e0	n/a	n/a	n/a	3.940	n/a
S961003898	A	Titanium - ICP-Acid Digest	ug/g	88.00	< 2.00e-01	< 78.60	< 7.76e1	n/a	n/a	n/a	78.60	n/a
S961003898	A	Thallium - ICP-Acid Digest	ug/g	90.70	< 5.00e-01	< 1.97e+02	< 1.94e2	n/a	n/a	n/a	197.0	n/a
S961003898	A	Uranium - ICP-Acid Digest	ug/g	91.80	< 5.00e-02	< 19.70	< 1.94e1	n/a	n/a	n/a	39.40	n/a
S961003898	A	Vanadium - ICP-Acid Digest	ug/g	88.80	1.60e-02	36.10	35.70	35.90	5.69	n/a	3.940	n/a
S961003898	A	Zinc - ICP-Acid Digest	ug/g	93.20	< 1.00e-02	4.440	4.700	4.570	5.69	n/a	3.940	n/a
S961003907	A	Acetate by IC-DIONE 4000/4500	ug/g	100.8	< 2.00e-02	8.56e+02	782.0	819.0	9.04	n/a	163.0	n/a
S961003907	A	Formate by IC-DIONE 4000/4500	ug/g	100.8	< 2.10e-02	1.36e+03	1.25e+03	1.30e+03	8.43	n/a	171.1	n/a
S961003907	A	Fluoride by IC-DIONE 4000/4500	ug/g	96.95	< 1.00e-02	7.77e+02	239.0	508.1	1.06	n/a	54.24	n/a
S961003907	A	Chloride by IC-DIONE 4000/4500	ug/g	99.49	< 1.07e-02	5.55e+03	5.71e+03	5.63e+03	2.84	n/a	70.94	n/a
S961003907	A	Nitrate by IC-DIONE 4000/4500	ug/g	95.77	< 1.70e-01	7.46e+04	7.71e+04	7.58e+04	3.30	n/a	446.5	n/a
S961003907	A	Nitrite by Ion Chromatograph	ug/g	97.45	< 1.28e-01	6.58e+02	724.0	690.9	9.55	n/a	525.8	n/a
S961003907	A	Phosphate by IC-DIONE 4000/4500	ug/g	97.98	1.88e-01	1.06e+05	1.06e+05	1.06e+05	0.00	n/a	583.8	n/a
S961003907	A	Sulfate by IC-DIONE 4000/4500	ug/g	105.5	< 1.19e-01	4.42e+03	1.55e+03	1.49e+03	8.75	n/a	496.2	n/a
S961003907	A	Sulfate by IC-DIONE 4000/4500	ug/g	97.94	< 1.36e-01	4.51e+03	4.68e+03	4.60e+03	3.70	n/a	467.2	n/a
S961003907	A	Oxalate by IC-DIONE 4000/4500	ug/g	98.49	< 1.03e-01	8.00e+03	7.92e+03	7.96e+03	1.01	n/a	436.3	n/a

L Lower Half of Segment: L Lower Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec.	Det Limit	Count Err%
S96T003817		DSC Exotherm using Mettler	Joules/g	96.31	n/a	16.10	7.600	11.95	71.7	n/a	n/a	n/a
S96T003817		DSC Exotherm DRY, Calculated	Joules/g	n/a	n/a	28.90	13.70	21.30	71.4	n/a	n/a	n/a
S96T003817		% Water by IGA using Mettler	%	99.48	n/a	44.82	43.89	44.36	2.10	n/a	n/a	n/a
S96T003827	F	Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.01e+03	<2.06e2	89.20	3.48	97.00	201.0	n/a
S96T003827	F	Aluminum - ICP-Fusion	ug/g	98.20	<1.00e-02	2.54e+04	2.63e+04	2.58e+04	3.48	97.00	1.00e+03	n/a
S96T003827	F	Arsenic - ICP-Fusion	ug/g	103.4	<1.00e-01	<2.01e+03	<2.06e2	102.0	n/a	102.0	2.01e+03	n/a
S96T003827	F	Boron - ICP-Fusion	ug/g	100.0	<1.00e-02	<1.01e+03	<1.03e3	n/a	n/a	95.20	1.00e+03	n/a
S96T003827	F	Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	93.70	1.00e+03	n/a
S96T003827	F	Beryllium - ICP-Fusion	ug/g	103.6	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	98.00	100.0	n/a
S96T003827	F	Bismuth - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	98.90	2.01e+03	n/a
S96T003827	F	Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	97.20	100.0	n/a
S96T003827	F	Cadmium - ICP-Fusion	ug/g	100.8	<5.00e-03	<1.01e+03	<1.03e2	n/a	n/a	94.90	2.01e+03	n/a
S96T003827	F	Cerium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	96.60	403.0	n/a
S96T003827	F	Cobalt - ICP-Fusion	ug/g	100.8	<1.00e-02	4.40e+02	4.42e2	1.22e+03	5.91	98.00	201.0	n/a
S96T003827	F	Chromium - ICP-Fusion	ug/g	101.4	<1.00e-02	1.15e+03	1.22e+03	1.18e+03	n/a	97.50	201.0	n/a
S96T003827	F	Copper - ICP-Fusion	ug/g	103.4	<1.00e-02	<2.01e+03	<2.06e2	n/a	n/a	101.0	1.00e+03	n/a
S96T003827	F	Iron - ICP-Fusion	ug/g	101.0	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	n/a	1.00e+04	n/a
S96T003827	F	Potassium - ICP-Fusion	ug/g	96.80	<5.00e-01	8.66e+06	n/a	n/a	n/a	97.00	1.00e+03	n/a
S96T003827	F	Lanthanum - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	97.00	1.00e+03	n/a
S96T003827	F	Lithium - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	95.30	201.0	n/a
S96T003827	F	Magnesium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	91.50	2.01e+03	n/a
S96T003827	F	Manganese - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.01e+03	<2.06e2	n/a	n/a	92.50	201.0	n/a
S96T003827	F	Molybdenum - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	97.00	1.00e+03	n/a
S96T003827	F	Sodium - ICP-Fusion	ug/g	98.80	1.700	2.35e+05	2.22e+05	2.28e+05	5.69	129.0	2.01e+03	n/a
S96T003827	F	Neodymium - ICP-Fusion	ug/g	101.8	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	96.40	2.01e+03	n/a
S96T003827	F	Nickel - ICP-Fusion	ug/g	100.4	1.24e-01	2.31e+03	4.42e2	n/a	n/a	96.70	403.0	n/a
S96T003827	F	Phosphorus - ICP-Fusion	ug/g	101.6	<2.00e-01	4.40e+03	4.42e2	n/a	n/a	101.0	4.03e+03	n/a
S96T003827	F	Lead - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	95.20	2.01e+03	n/a
S96T003827	F	Sulfur - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	106.0	2.01e+03	n/a
S96T003827	F	Antimony - ICP-Fusion	ug/g	95.60	<5.00e-02	<1.21e+03	<1.24e3	n/a	n/a	92.30	1.21e+03	n/a
S96T003827	F	Selenium - ICP-Fusion	ug/g	97.80	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	96.70	2.01e+03	n/a
S96T003827	F	Samarium - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	95.80	1.00e+03	n/a
S96T003827	F	Strontium - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.01e+03	<2.06e2	n/a	n/a	93.30	201.0	n/a
S96T003827	F	Titanium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.01e+03	<2.06e2	n/a	n/a	92.30	4.03e+03	n/a
S96T003827	F	Thallium - ICP-Fusion	ug/g	97.40	<1.00e-02	<2.01e+03	<2.06e2	n/a	n/a	86.50	1.00e+03	n/a
S96T003827	F	Thallium - ICP-Fusion	ug/g	96.80	<5.00e-01	4.40e+03	4.42e3	n/a	n/a	96.10	1.00e+03	n/a
S96T003827	F	Uranium - ICP-Fusion	ug/g	96.00	<5.00e-01	<1.01e+03	<1.03e3	n/a	n/a	98.50	2.01e+03	n/a
S96T003827	F	Vanadium - ICP-Fusion	ug/g	100.2	<5.00e-02	<1.01e+03	<1.03e3	n/a	n/a	96.10	1.00e+03	n/a
S96T003827	F	Zinc - ICP-Fusion	ug/g	101.2	1.10e-02	4.32e+02	810.0	621.0	60.9	98.50	201.0	n/a
S96T003827	F	Zirconium - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.01e+03	<2.06e2	n/a	n/a	94.10	201.0	n/a
S96T003827	F	Alpha of Digested Solid	ug/g	84.38	<1.46e-02	<2.67e-02	<1.52E-2	13.85	5.05	n/a	3.70e-02	5.00e+02
S96T003828	A	Aluminum - ICP-Acid Digest	ug/g	23.40	<1.00e-02	13.50	14.20	2.56e+04	1.17	n/a	4.040	n/a
S96T003828	A	Aluminum - ICP-Acid Digest	ug/g	101.4	2.28e-01	2.55e+04	2.58e+04	2.56e+04	1.77	n/a	20.20	n/a
S96T003828	A	Arsenic - ICP-Acid Digest	ug/g	95.00	<1.00e-01	<40.40	<3.93e1	n/a	n/a	40.40	40.40	n/a
S96T003828	A	Boron - ICP-Acid Digest	ug/g	118.2	6.30e-01	88.70	90.00	89.35	1.45	n/a	20.20	n/a
S96T003828	A	Barium - ICP-Acid Digest	ug/g	92.60	<5.00e-02	<20.20	<1.97e1	n/a	n/a	20.20	20.20	n/a
S96T003828	A	Beryllium - ICP-Acid Digest	ug/g	94.80	<5.00e-03	<2.020	<1.97e0	n/a	n/a	n/a	2.020	n/a
S96T003828	A	Bismuth - ICP-Acid Digest	ug/g	92.40	<1.00e-01	<40.40	<3.93e1	n/a	n/a	n/a	40.40	n/a

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003828	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	1.30e+02	161.0	145.5	21.3	N/A	40.40	N/A
S961003828	A	Cadmium - ICP-Acid Digest	ug/g	90.80	<1.00e-03	4.820	5.900	5.360	20.1	N/A	2.020	N/A
S961003828	A	Cerium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 4.00	<3.93e1	N/A	N/A	N/A	4.040	N/A
S961003828	A	Cobalt - ICP-Acid Digest	ug/g	93.80	<2.00e-02	< 8.080	<7.86e0	N/A	N/A	N/A	8.080	N/A
S961003828	A	Chromium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	1.21e+03	1.22e+03	1.22e+03	0.82	N/A	4.040	N/A
S961003828	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	7.150	6.240	6.695	13.6	N/A	4.040	N/A
S961003828	A	Iron - ICP-Acid Digest	ug/g	93.00	<5.00e-02	46.10	47.10	46.60	2.15	N/A	20.20	N/A
S961003828	A	Potassium - ICP-Acid Digest	ug/g	93.80	<5.00e-01	3.95e+03	4.02e+03	3.98e+03	1.76	N/A	202.0	N/A
S961003828	A	Lanthanum - ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 20.20	<1.97e1	N/A	N/A	N/A	20.20	N/A
S961003828	A	Lithium - ICP-Acid Digest	ug/g	96.60	<1.00e-02	12.80	13.50	13.15	5.32	N/A	4.040	N/A
S961003828	A	Magnesium - ICP-Acid Digest	ug/g	88.60	<1.00e-01	< 4.00	<3.93e1	N/A	N/A	N/A	4.040	N/A
S961003828	A	Manganese - ICP-Acid Digest	ug/g	90.00	<5.00e-02	< 4.040	<3.93e0	N/A	N/A	N/A	4.040	N/A
S961003828	A	Molybdenum - ICP-Acid Digest	ug/g	94.60	<5.00e-02	59.50	59.40	59.45	0.17	N/A	20.20	N/A
S961003828	A	Sodium - ICP-Acid Digest	ug/g	150.6	1.900	1.83e+05	1.85e+05	1.84e+05	1.09	N/A	40.40	N/A
S961003828	A	Niobium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	< 4.00	<3.93e1	N/A	N/A	N/A	40.40	N/A
S961003828	A	Nickel - ICP-Acid Digest	ug/g	93.40	<2.00e-02	20.50	19.90	20.20	2.87	N/A	8.080	N/A
S961003828	A	Phosphorus - ICP-Acid Digest	ug/g	89.00	<1.00e-01	3.89e+02	382.0	385.5	1.92	N/A	40.40	N/A
S961003828	A	Lead - ICP-Acid Digest	ug/g	88.40	<6.00e-02	< 4.00	<3.93e1	N/A	N/A	N/A	40.40	N/A
S961003828	A	Antimony - ICP-Acid Digest	ug/g	93.00	1.19e-01	2.18e+03	2.23e+03	2.20e+03	2.27	N/A	40.40	N/A
S961003828	A	Selenium - ICP-Acid Digest	ug/g	87.40	<1.00e-02	< 24.20	<2.36e1	N/A	N/A	N/A	24.20	N/A
S961003828	A	Strontium - ICP-Acid Digest	ug/g	244.0	5.31e-01	1.65e+02	181.0	173.0	9.25	N/A	40.40	N/A
S961003828	A	Titanium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 4.00	<3.93e1	N/A	N/A	N/A	40.40	N/A
S961003828	A	Tantalum - ICP-Acid Digest	ug/g	92.40	<1.00e-02	< 4.040	<3.93e0	N/A	N/A	N/A	4.040	N/A
S961003828	A	Thallium - ICP-Acid Digest	ug/g	88.00	<2.00e-01	< 80.80	<7.86e1	N/A	N/A	N/A	80.80	N/A
S961003828	A	Uranium - ICP-Acid Digest	ug/g	90.70	<5.00e-01	<2.02e+02	<1.97e2	N/A	N/A	N/A	202.0	N/A
S961003828	A	Vanadium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	< 20.20	<1.97e1	N/A	N/A	N/A	20.20	N/A
S961003828	A	Zinc - ICP-Acid Digest	ug/g	88.80	1.60e-02	35.90	37.40	36.65	4.09	N/A	4.040	N/A
S961003828	A	Zirconium - ICP-Acid Digest	ug/g	93.20	<1.00e-02	4.440	4.150	4.295	6.75	N/A	4.040	N/A
S961003829	A	Acetate by IC-DIONEX 4000/4500	ug/g	102.4	<2.00e-02	6.55e+02	784.0	709.7	15.4	N/A	162.6	N/A
S961003829	A	Formate by IC-DIONEX 4000/4500	ug/g	100.0	<2.00e-02	1.02e+03	1.17e+03	1.10e+03	13.7	N/A	170.7	N/A
S961003829	A	Fluoride by IC-DIONEX 4000/4500	ug/g	98.64	<1.00e-02	4.12e+02	455.0	433.4	9.92	N/A	28.35	N/A
S961003829	A	Chloride by IC-DIONEX 4000/4500	ug/g	104.3	2.00e-02	3.76e+03	4.05e+03	3.91e+03	7.43	N/A	37.07	N/A
S961003829	A	Nitrite by IC-DIONEX 4000/4500	ug/g	100.2	<1.07e-01	4.74e+04	5.23e+04	4.98e+04	9.83	N/A	233.3	N/A
S961003829	A	Bromide by Ion Chromatograph	ug/g	105.3	<1.26e-01	<2.75e+02	340.0	N/A	N/A	N/A	274.8	N/A
S961003829	A	Nitrate by IC-DIONEX 4000/4500	ug/g	102.4	<1.00e-01	6.05e+04	6.56e+04	6.31e+04	8.09	N/A	305.1	N/A
S961003829	A	Phosphate by IC-DIONEX 4000/4500	ug/g	101.8	<1.19e-01	<2.59e+02	524.0	N/A	N/A	N/A	259.3	N/A
S961003829	A	Sulfate by IC-DIONEX 4000/4500	ug/g	104.9	<1.36e-01	3.33e+03	3.51e+03	3.42e+03	5.26	N/A	296.3	N/A
S961003829	A	Oxalate by IC-DIONEX 4000/4500	ug/g	104.5	<1.05e-01	4.78e+03	5.40e+03	5.09e+03	12.2	N/A	229.0	N/A

Drainable Liquid; Drainable Liquid

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003751	D	DSC Exotherm Using Mettler	Joules/g	102.6	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S961003751	D	DSC Exotherm Dry Calculated	Joules/g Dry	N/A	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S961003751	D	Specific Gravity	SP. G.	98.74	N/A	1.400	1.395	1.397	0.36	N/A	1.00e-03	N/A
S961003751	D	% Water by TGA on Perkin Elmer	%	99.31	N/A	51.31	50.95	51.13	16.80	N/A	4.010	N/A
S961003751	D	Diluter by ICP-Acid Dil.	ug/mL	100.0	<1.00e-02	17.10	16.50	16.80	3.57	89.20	4.010	N/A
S961003751	D	Aluminum-ICP-Acid Dil.	ug/mL	98.00	<5.00e-02	3.87e+04	3.91e+04	3.89e+04	1.03	-1.190e2	20.10	N/A

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003751	D	Arsenic-ICP-Acid Dill.	ug/mL	102.8	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	88.90	40.10	n/a
S961003751	D	Boron-ICP-Acid Dill.	ug/mL	99.40	<5.00e-02	75.60	75.00	75.30	0.80	95.90	20.10	n/a
S961003751	D	Barium-ICP-Acid Dill.	ug/mL	100.0	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	95.80	20.10	n/a
S961003751	D	Beryllium-ICP-Acid Dill.	ug/mL	102.2	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	98.90	2.000	n/a
S961003751	D	Bismuth-ICP-Acid Dill.	ug/mL	101.8	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.40	40.10	n/a
S961003751	D	Calcium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	104.0	40.10	n/a
S961003751	D	Cadmium-ICP-Acid Dill.	ug/mL	100.8	<5.00e-03	< 2.000	<2.00e0	n/a	n/a	98.70	2.000	n/a
S961003751	D	Cerium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.40	40.10	n/a
S961003751	D	Cobalt-ICP-Acid Dill.	ug/mL	101.4	<1.00e-02	< 8.020	<8.02e0	n/a	n/a	98.70	8.020	n/a
S961003751	D	Chromium-ICP-Acid Dill.	ug/mL	101.4	<1.00e-02	< 1.96e+02	<195.0	195.5	0.51	98.10	4.010	n/a
S961003751	D	Copper-ICP-Acid Dill.	ug/mL	103.4	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	99.30	4.010	n/a
S961003751	D	Iron-ICP-Acid Dill.	ug/mL	101.0	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	99.90	20.10	n/a
S961003751	D	Potassium-ICP-Acid Dill.	ug/mL	96.60	<5.00e-01	< 6.07e+03	<6.13e+03	6.13e+03	1.96	73.30	200.0	n/a
S961003751	D	Lanthanum-ICP-Acid Dill.	ug/mL	102.2	<1.00e-02	< 20.10	<2.01e1	n/a	n/a	100.0	20.10	n/a
S961003751	D	Lithium-ICP-Acid Dill.	ug/mL	101.6	<1.00e-02	< 7.560	<7.56e0	n/a	n/a	92.80	40.10	n/a
S961003751	D	Magnesium-ICP-Acid Dill.	ug/mL	97.40	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	94.40	40.10	n/a
S961003751	D	Manganese-ICP-Acid Dill.	ug/mL	98.40	<1.00e-01	< 4.010	<4.01e0	n/a	n/a	99.20	40.10	n/a
S961003751	D	Molybdenum-ICP-Acid Dill.	ug/mL	100.8	<1.00e-01	< 87.60	<88.50	88.05	1.02	99.20	40.10	n/a
S961003751	D	Sodium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-01	< 2.34e+05	<2.34e+05	2.34e+05	0.43	-1.590e3	40.10	n/a
S961003751	D	Neodymium-ICP-Acid Dill.	ug/mL	102.6	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	100.0	40.10	n/a
S961003751	D	Nickel-ICP-Acid Dill.	ug/mL	100.2	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	96.90	8.020	n/a
S961003751	D	Phosphorus-ICP-Acid Dill.	ug/mL	101.8	<1.00e-01	< 5.28e+02	<384.0	456.0	31.6	49.40	80.20	n/a
S961003751	D	Lead-ICP-Acid Dill.	ug/mL	100.6	<1.00e-01	< 50.90	<43.60	47.25	15.4	99.70	40.10	n/a
S961003751	D	Sulfur-ICP-Acid Dill.	ug/mL	99.40	<1.00e-01	< 9.64e+02	<977.0	970.5	1.34	95.10	40.10	n/a
S961003751	D	Selenium-ICP-Acid Dill.	ug/mL	96.20	<6.00e-02	< 24.10	<2.41e1	n/a	n/a	95.10	24.10	n/a
S961003751	D	Silicon-ICP-Acid Dill.	ug/mL	97.20	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	96.90	40.10	n/a
S961003751	D	Strontium-ICP-Acid Dill.	ug/mL	99.60	<5.00e-02	< 2.42e+02	<231.0	236.5	4.65	95.20	20.10	n/a
S961003751	D	Tantalum-ICP-Acid Dill.	ug/mL	98.80	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	96.80	40.10	n/a
S961003751	D	Titanium-ICP-Acid Dill.	ug/mL	99.60	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	96.70	4.010	n/a
S961003751	D	Thallium-ICP-Acid Dill.	ug/mL	98.00	<2.00e-01	< 80.20	<8.02e1	n/a	n/a	95.40	80.20	n/a
S961003751	D	Uranium-ICP-Acid Dill.	ug/mL	96.40	<5.00e-01	<2.00e+02	<2.00e2	n/a	n/a	96.00	200.0	n/a
S961003751	D	Vanadium-ICP-Acid Dill.	ug/mL	100.6	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	98.40	20.10	n/a
S961003751	D	Zinc-ICP-Acid Dill.	ug/mL	101.4	<1.00e-02	< 45.40	<40.90	43.15	10.4	100.0	4.010	n/a
S961003751	D	Zirconium-ICP-Acid Dill.	ug/mL	99.20	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	98.50	4.010	n/a
S961003751	D	Acetate by IC-Dionex 4000/4500	ug/mL	112.1	<2.00e-02	< 7.00e+02	<757.0	728.7	7.82	84.68	103.0	n/a
S961003751	D	Formate by Ion Chromatograph	ug/mL	99.17	<2.00e-02	< 1.00e+03	<1.75e+03	2.90	103.3	108.2	103.0	n/a
S961003751	D	Fluoride-IC-Dionex 4000/4500	ug/mL	98.64	<1.30e-02	< 53.83	<5.38e1	n/a	n/a	41.53	53.83	n/a
S961003751	D	Chloride-IC-Dionex 4000/4500	ug/mL	102.3	<1.70e-02	< 1.06e+04	<1.06e+04	1.06e+04	0.95	57.22	70.40	n/a
S961003751	D	Nitrite-IC - Dionex 4000/4500	ug/mL	97.43	<1.07e-01	< 1.19e+05	<1.20e+05	1.20e+05	0.84	91.91	443.1	n/a
S961003751	D	Bromide by Ion Chromatograph	ug/mL	96.94	<1.28e-01	< 1.66e+03	<1.66e+03	1.66e+03	0.00	93.72	521.8	n/a
S961003751	D	Nitrate by IC-Dionex 4000/4500	ug/mL	95.45	<1.40e-01	< 1.59e+05	<1.59e+05	1.59e+05	0.00	89.73	599.7	n/a
S961003751	D	Phosphate-IC-Dionex 4000/4500	ug/mL	106.4	<1.19e-01	< 1.33e+03	<4.92e2	n/a	n/a	102.0	492.4	n/a
S961003751	D	Sulfate by IC-Dionex 4000/4500	ug/mL	96.04	<1.36e-01	< 1.38e+03	<1.41e+03	1.39e+03	2.15	93.98	562.8	n/a
S961003751	D	Oxalate by IC-Dionex 4000/4500	ug/mL	99.43	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	99.24	434.8	n/a
S961003751	D	Alpha in Liquid Samples	uCi/mL	93.75	<5.59e-02	<7.97e-02	<6.49e-2	n/a	n/a	n/a	9.70e-02	5.00e+02

Table 1. Data Summary Table
AM-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961003706		Bulk Density of Sample	g/mL	n/a	n/a	1.590	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S961003792		DSC Exotherm using Mettler	Joules/g	104.0	n/a	31.70	31.70	31.0	4.2	n/a	n/a	n/a	n/a
S961003792		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	43.40	67.30	55.35	43.2	n/a	n/a	n/a	n/a
S961003792		% Water by TGA using Mettler	%	101.0	n/a	44.31	43.65	43.97	1.55	n/a	n/a	n/a	n/a
S961003895		F Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	82.80	205.0	n/a	n/a
S961003895		F Aluminum - ICP-Fusion	ug/g	98.20	<5.00e-02	2.43e+04	2.40e+04	2.42e+04	1.24	93.20	1.02e+03	n/a	n/a
S961003895		F Arsenic - ICP-Fusion	ug/g	103.4	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	103.0	2.05e+03	n/a	n/a
S961003895		F Boron - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	94.90	1.02e+03	n/a	n/a
S961003895		F Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	93.60	1.02e+03	n/a	n/a
S961003895		F Beryllium - ICP-Fusion	ug/g	103.6	<5.00e-03	<1.02e+02	<1.00e2	n/a	n/a	97.90	102.0	n/a	n/a
S961003895		F Bismuth - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	97.90	2.05e+03	n/a	n/a
S961003895		F Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	101.0	2.05e+03	n/a	n/a
S961003895		F Cadmium - ICP-Fusion	ug/g	100.0	<5.00e-03	<1.02e+02	<1.00e2	n/a	n/a	95.70	102.0	n/a	n/a
S961003895		F Cerium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	97.70	2.05e+03	n/a	n/a
S961003895		F Cobalt - ICP-Fusion	ug/g	101.4	<1.00e-02	<2.05e+03	2.13e+03	2.09e+03	3.83	96.90	205.0	n/a	n/a
S961003895		F Chromium - ICP-Fusion	ug/g	103.4	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	96.80	205.0	n/a	n/a
S961003895		F Copper - ICP-Fusion	ug/g	101.0	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	98.30	1.02e+04	n/a	n/a
S961003895		F Iron - ICP-Fusion	ug/g	96.80	<5.00e-01	8.88e+06	n/a	n/a	n/a	n/a	1.02e+03	n/a	n/a
S961003895		F Potassium - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	97.30	1.02e+03	n/a	n/a
S961003895		F Lanthanum - ICP-Fusion	ug/g	102.0	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	94.20	205.0	n/a	n/a
S961003895		F Lithium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	92.80	2.05e+03	n/a	n/a
S961003895		F Magnesium - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	93.60	205.0	n/a	n/a
S961003895		F Manganese - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	97.90	1.02e+03	n/a	n/a
S961003895		F Molybdenum - ICP-Fusion	ug/g	98.80	<1.00e-01	2.52e+05	2.22e+05	2.36e+05	11.9	84.50	2.05e+03	n/a	n/a
S961003895		F Sodium - ICP-Fusion	ug/g	101.8	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	96.40	2.05e+03	n/a	n/a
S961003895		F Neodymium - ICP-Fusion	ug/g	100.4	1.24e-01	3.54e+03	1.38e+03	2.46e+03	87.8	98.90	409.0	n/a	n/a
S961003895		F Nickel - ICP-Fusion	ug/g	99.80	<2.00e-01	<2.04e+03	<2.01e3	n/a	n/a	97.10	2.05e+03	n/a	n/a
S961003895		F Phosphorus - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	97.80	2.05e+03	n/a	n/a
S961003895		F Lead - ICP-Fusion	ug/g	99.00	<1.00e-01	2.89e+03	2.88e+03	2.88e+03	0.35	97.80	2.05e+03	n/a	n/a
S961003895		F Sulfur - ICP-Fusion	ug/g	95.60	<6.00e-02	<1.23e+03	<1.21e3	n/a	n/a	92.60	1.23e+03	n/a	n/a
S961003895		F Antimony - ICP-Fusion	ug/g	97.80	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	96.30	2.05e+03	n/a	n/a
S961003895		F Selenium - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	94.90	2.05e+03	n/a	n/a
S961003895		F Silicon - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	94.30	205.0	n/a	n/a
S961003895		F Samarium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	93.50	205.0	n/a	n/a
S961003895		F Strontium - ICP-Fusion	ug/g	97.40	<1.00e-02	<4.09e+03	<4.02e3	n/a	n/a	93.90	4.09e+03	n/a	n/a
S961003895		F Tantalum - ICP-Fusion	ug/g	96.80	<2.00e-01	<1.02e+04	<1.00e4	n/a	n/a	89.50	1.02e+04	n/a	n/a
S961003895		F Thallium - ICP-Fusion	ug/g	100.2	<5.00e-02	<1.02e+03	<1.00e3	n/a	n/a	97.20	1.02e+03	n/a	n/a
S961003895		F Uranium - ICP-Fusion	ug/g	96.00	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	94.90	205.0	n/a	n/a
S961003895		F Vanadium - ICP-Fusion	ug/g	101.2	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	94.90	205.0	n/a	n/a
S961003895		F Zinc - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	94.90	205.0	n/a	n/a
S961003895		F Zirconium - ICP-Fusion	ug/g	23.40	<1.00e-02	14.80	15.10	14.95	2.01	n/a	3.960	n/a	n/a
S961003899		F Silver - ICP-Acid Digest	ug/g	101.4	2.28e-01	2.33e+04	2.39e+04	2.36e+04	2.54	n/a	19.80	n/a	n/a
S961003899		A Aluminum - ICP-Acid Digest	ug/g										

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003899	A	Arsenic - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Boron - ICP-Acid Digest	ug/g	118.2	6.30e-01	59.30	67.70	63.50	13.2	n/a	19.80	n/a
S961003899	A	Barium - ICP-Acid Digest	ug/g	92.80	<5.00e-02	< 19.80	<1.99e1	n/a	n/a	n/a	19.80	n/a
S961003899	A	Beryllium - ICP-Acid Digest	ug/g	94.60	<5.00e-03	< 1.980	<1.99e1	n/a	n/a	n/a	1.980	n/a
S961003899	A	Bismuth - ICP-Acid Digest	ug/g	92.40	<1.00e-01	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	3.23e+02	280.0	301.5	14.3	n/a	39.60	n/a
S961003899	A	Cadmium - ICP-Acid Digest	ug/g	90.80	<5.00e-03	9.500	9.610	9.555	1.15	n/a	1.980	n/a
S961003899	A	Cerium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Cobalt - ICP-Acid Digest	ug/g	93.80	<2.00e-02	< 7.920	<7.96e0	n/a	n/a	n/a	7.920	n/a
S961003899	A	Chromium - ICP-Acid Digest	ug/g	92.20	1.30e-02	2.02e+03	2.07e+03	2.04e+03	2.44	n/a	3.960	n/a
S961003899	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	6.280	6.480	6.380	3.13	n/a	3.960	n/a
S961003899	A	Iron - ICP-Acid Digest	ug/g	93.00	<5.00e-02	72.60	327.0	199.8	127	n/a	19.80	n/a
S961003899	A	Potassium - ICP-Acid Digest	ug/g	93.80	<5.00e-01	3.61e+03	3.68e+03	3.68e+03	4.07	n/a	198.0	n/a
S961003899	A	Lithium - ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 19.80	<1.99e1	n/a	n/a	n/a	19.80	n/a
S961003899	A	Lanthanum - ICP-Acid Digest	ug/g	96.60	<1.00e-02	8.940	7.130	8.035	22.5	n/a	3.960	n/a
S961003899	A	Magnesium - ICP-Acid Digest	ug/g	88.60	<1.00e-02	8.400	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Manganese - ICP-Acid Digest	ug/g	90.00	<1.00e-02	8.400	10.30	9.350	20.3	n/a	3.960	n/a
S961003899	A	Molybdenum - ICP-Acid Digest	ug/g	94.60	<5.00e-02	55.80	55.40	55.60	0.72	n/a	19.80	n/a
S961003899	A	Sodium - ICP-Acid Digest	ug/g	150.6	1.900	1.87e+05	1.92e+05	1.90e+05	2.64	n/a	39.60	n/a
S961003899	A	Neodymium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Nickel - ICP-Acid Digest	ug/g	93.40	<2.00e-02	34.30	37.90	36.10	9.97	n/a	7.920	n/a
S961003899	A	Phosphorus - ICP-Acid Digest	ug/g	93.40	<2.00e-01	6.53e+02	510.0	581.5	24.6	n/a	79.20	n/a
S961003899	A	Lead - ICP-Acid Digest	ug/g	89.00	<1.00e-01	41.00	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Sulfur - ICP-Acid Digest	ug/g	93.00	1.19e-01	2.94e+03	2.99e+03	2.96e+03	1.69	n/a	39.60	n/a
S961003899	A	Antimony - ICP-Acid Digest	ug/g	88.40	<5.00e-02	< 23.80	<2.39e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Selenium - ICP-Acid Digest	ug/g	87.40	<1.00e-01	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Silicon - ICP-Acid Digest	ug/g	244.0	5.31e-01	1.34e+02	142.0	138.0	5.80	n/a	19.80	n/a
S961003899	A	Samarium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 39.60	<3.98e1	n/a	n/a	n/a	39.60	n/a
S961003899	A	Strontium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.960	<3.98e0	n/a	n/a	n/a	3.960	n/a
S961003899	A	Titanium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.960	<3.98e0	n/a	n/a	n/a	3.960	n/a
S961003899	A	Thallium - ICP-Acid Digest	ug/g	88.00	<2.00e-01	< 79.30	<7.96e1	n/a	n/a	n/a	79.20	n/a
S961003899	A	Uranium - ICP-Acid Digest	ug/g	90.70	<5.00e-01	<1.99e+02	<1.99e2	n/a	n/a	n/a	198.0	n/a
S961003899	A	Vanadium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	< 19.80	<1.99e1	n/a	n/a	n/a	19.80	n/a
S961003899	A	Zinc - ICP-Acid Digest	ug/g	88.80	1.60e-02	35.90	35.40	35.65	1.40	n/a	3.960	n/a
S961003908	W	Zirconium - ICP-Acid Digest	ug/g	93.20	<1.00e-02	6.230	6.370	6.300	2.22	n/a	3.960	n/a
S961003908	W	Acetate by IC-Dionex 4000/4500	ug/g	102.4	<2.00e-02	<1.66e+02	<736.0	n/a	n/a	n/a	166.5	n/a
S961003908	W	Formate by IC-Dionex 4000/4500	ug/g	100.0	<2.00e-02	<1.29e+03	<1.25e+03	1.27e+03	3.15	n/a	174.8	n/a
S961003908	W	Fluoride-IC-Dionex 4000/4500	ug/g	98.64	1.30e-02	7.08e+02	728.0	718.1	2.79	n/a	29.03	n/a
S961003908	W	Chloride-IC-Dionex 4000/4500	ug/g	104.3	2.00e-02	4.16e+03	4.08e+03	4.12e+03	1.94	n/a	37.96	n/a
S961003908	W	Nitrite-IC - Dionex 4000/4500	ug/g	100.2	<1.07e-01	5.25e+04	5.29e+04	5.27e+04	0.76	n/a	238.9	n/a
S961003908	W	Bromide by Ion Chromatograph	ug/g	105.3	<1.26e-01	<2.81e+02	<2.70e2	n/a	n/a	n/a	281.4	n/a
S961003908	W	Nitrate by IC-Dionex 4000/4500	ug/g	102.4	<1.40e-01	6.94e+04	8.07e+04	7.51e+04	15.1	n/a	312.4	n/a
S961003908	W	Phosphate-IC-Dionex 4000/4500	ug/g	101.8	<1.19e-01	1.04e+03	1.08e+03	1.36e+03	47.1	n/a	265.5	n/a
S961003908	W	Sulfate by IC-Dionex 4000/4500	ug/g	104.9	<1.36e-01	5.72e+03	5.82e+03	5.77e+03	1.73	n/a	303.5	n/a
S961003908	W	Oxalate by IC-Dionex 4000/4500	ug/g	104.5	<1.09e-01	5.47e+03	5.66e+03	5.57e+03	3.41	n/a	234.5	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

L Lower Half of Segment: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003705		Bulk Density of Sample	g/mL	n/a	n/a	1.580	n/a	n/a	n/a	n/a	5.00e-01	n/a
S96T003792		DSC Exotherm using Mettler	Joules/g	104.0	n/a	27.20	28.00	27.60	2.90	n/a	n/a	n/a
S96T003793		DSC Exotherm DRY Calculated	Joules/g Dry	n/a	n/a	48.00	49.40	48.70	2.87	n/a	n/a	n/a
S96T003793		% Water by TGA using Mettler	%	101.0	n/a	42.04	44.58	43.31	5.86	n/a	n/a	n/a
S96T003889		F Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.03e-02	<2.00e2	2.42e+04	2.45	88.90	203.0	n/a
S96T003889		F Arsenic - ICP-Fusion	ug/g	98.80	<5.00e-02	<2.48e-04	<2.00e3	2.42e+04	n/a	94.60	1.01e+03	n/a
S96T003889		F Boron - ICP-Fusion	ug/g	104.4	<1.00e-01	<2.03e-03	<2.00e3	n/a	n/a	95.00	1.01e+03	n/a
S96T003889		F Barium - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.01e-03	<1.00e3	n/a	n/a	94.10	1.01e+03	n/a
S96T003889		F Beryllium - ICP-Fusion	ug/g	103.8	<5.00e-01	<2.03e-03	<2.00e2	n/a	n/a	98.00	101.0	n/a
S96T003889		F Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.03e-03	<2.00e3	n/a	n/a	99.30	2.03e+03	n/a
S96T003889		F Cadmium - ICP-Fusion	ug/g	101.4	<5.00e-03	<2.03e-03	<2.00e2	n/a	n/a	97.50	101.0	n/a
S96T003889		F Calcium - ICP-Fusion	ug/g	100.8	<1.00e-01	<2.03e-03	<2.00e3	n/a	n/a	98.30	2.03e+03	n/a
S96T003889		F Cerium - ICP-Fusion	ug/g	101.4	<1.00e-02	<2.03e-03	<2.00e2	n/a	n/a	95.40	2.03e+03	n/a
S96T003889		F Cobalt - ICP-Fusion	ug/g	101.8	<1.00e-02	<4.06e-02	<4.01e2	n/a	n/a	97.40	406.0	n/a
S96T003889		F Chromium - ICP-Fusion	ug/g	104.2	<1.00e-02	<2.03e-02	<1.75e+03	1.72e+03	2.90	97.60	203.0	n/a
S96T003889		F Copper - ICP-Fusion	ug/g	101.6	<5.00e-02	<2.03e-02	<1.01e+03	n/a	n/a	98.40	203.0	n/a
S96T003889		F Iron - ICP-Fusion	ug/g	100.0	<5.00e-01	<6.07e+06	<1.00e3	n/a	n/a	97.30	1.01e+04	n/a
S96T003889		F Potassium - ICP-Fusion	ug/g	102.4	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	97.00	1.01e+03	n/a
S96T003889		F Lanthanum - ICP-Fusion	ug/g	102.8	<1.00e-02	<2.03e-02	<2.00e2	n/a	n/a	96.50	203.0	n/a
S96T003889		F Lithium - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.03e+03	<2.00e2	n/a	n/a	93.60	2.03e+03	n/a
S96T003889		F Magnesium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.03e+03	<2.00e2	n/a	n/a	92.90	203.0	n/a
S96T003889		F Manganese - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	99.40	1.01e+03	n/a
S96T003889		F Molybdenum - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.03e+03	<2.00e2	n/a	n/a	97.10	1.01e+03	n/a
S96T003889		F Sodium - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.03e+03	<2.00e2	2.30e+05	4.35	99.40	2.03e+03	n/a
S96T003889		F Neodymium - ICP-Fusion	ug/g	101.4	8.50e-02	<4.06e-02	<1.32e+03	n/a	n/a	97.30	2.03e+03	n/a
S96T003889		F Nickel - ICP-Fusion	ug/g	102.0	<2.00e-01	<4.06e+03	<4.01e3	n/a	n/a	99.90	406.0	n/a
S96T003889		F Phosphorus - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.03e+03	<2.00e3	n/a	n/a	96.70	2.03e+03	n/a
S96T003889		F Lead - ICP-Fusion	ug/g	100.0	<1.00e-01	<3.00e+03	<3.14e+03	3.07e+03	4.56	95.30	2.03e+03	n/a
S96T003889		F Sulfur - ICP-Fusion	ug/g	96.00	<6.00e-02	<1.22e+03	<1.20e3	n/a	n/a	92.80	1.22e+03	n/a
S96T003889		F Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.03e+03	<2.00e3	n/a	n/a	95.20	2.03e+03	n/a
S96T003889		F Selenium - ICP-Fusion	ug/g	95.40	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	94.10	1.01e+03	n/a
S96T003889		F Silicon - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.03e+03	<2.00e3	n/a	n/a	92.60	2.03e+03	n/a
S96T003889		F Strontium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.03e+02	<2.00e2	n/a	n/a	94.50	203.0	n/a
S96T003889		F Titanium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.03e+02	<2.00e2	n/a	n/a	92.90	203.0	n/a
S96T003889		F Thallium - ICP-Fusion	ug/g	96.70	<2.00e-01	<4.06e+03	<4.01e3	n/a	n/a	92.70	4.06e+03	n/a
S96T003889		F Uranium - ICP-Fusion	ug/g	100.6	<5.00e-02	<1.01e+04	<1.00e3	n/a	n/a	88.00	1.01e+04	n/a
S96T003889		F Vanadium - ICP-Fusion	ug/g	101.4	<5.00e-02	<2.03e+02	<2.00e2	n/a	n/a	96.10	1.01e+04	n/a
S96T003889		F Zinc - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.03e+02	<2.00e2	n/a	n/a	94.10	203.0	n/a
S96T003889		F Zirconium - ICP-Fusion	ug/g	97.66	<2.72e-02	<3.16e-02	<2.43e-2	n/a	n/a	85.24	4.70e-02	5.00e+02
S96T003890		A Alpha of Digested Solid	ug/g	23.40	<1.00e-02	<1.70e-02	<2.47e-04	14.70	0.00	60.50	3.920	n/a
S96T003900		A Aluminum - ICP-Acid Digest	ug/g	101.4	<2.28e-01	<2.41e+04	<3.93e1	2.44e+04	2.46	112.6	19.60	n/a
S96T003900		A Arsenic - ICP-Acid Digest	ug/g	95.00	<1.00e-01	<3.90	<3.93e1	n/a	n/a	96.40	39.20	n/a
S96T003900		A Boron - ICP-Acid Digest	ug/g	118.2	<6.30e-01	<5.60	76.70	67.65	26.8	100.8	19.60	n/a
S96T003900		A Barium - ICP-Acid Digest	ug/g	92.80	<5.00e-02	<1.96e1	<1.96e1	n/a	n/a	95.60	19.60	n/a
S96T003900		A Beryllium - ICP-Acid Digest	ug/g	94.60	<5.00e-03	<1.96e0	<1.96e0	n/a	n/a	95.60	1.960	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err-%
S961003590	A	Bismuth - ICP-Acid Digest	ug/g	92.40	<1.00e-01	< 39.30	<3.93e1	n/a	96.40	39.20	n/a	n/a
S961003590	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	2.79e+02	536.0	63.1	93.40	39.20	n/a	n/a
S961003590	A	Calcium - ICP-Acid Digest	ug/g	90.80	<5.00e-03	10.80	11.60	11.20	7.14	1.960	n/a	n/a
S961003590	A	Cerium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 39.30	<3.93e1	n/a	98.40	39.20	n/a	n/a
S961003590	A	Cobalt - ICP-Acid Digest	ug/g	93.80	<2.00e-02	< 7.860	<7.86e0	n/a	96.20	7.860	n/a	n/a
S961003590	A	Chromium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	1.71e+03	1.75e+03	2.31	97.40	3.920	n/a	n/a
S961003590	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	5.350	6.030	5.690	12.0	93.80	n/a	n/a
S961003590	A	Iron - ICP-Acid Digest	ug/g	93.00	<5.00e-02	90.40	80.00	85.20	92.00	19.60	n/a	n/a
S961003590	A	Potassium - ICP-Acid Digest	ug/g	93.80	<5.00e-01	3.77e+03	3.77e+03	0.00	99.00	196.0	n/a	n/a
S961003590	A	Lanthanum - ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 19.60	<1.96e1	n/a	99.00	19.60	n/a	n/a
S961003590	A	Lithium - ICP-Acid Digest	ug/g	96.60	<1.00e-02	< 3.930	<3.93e0	n/a	101.6	3.920	n/a	n/a
S961003590	A	Magnesium - ICP-Acid Digest	ug/g	88.60	<1.00e-01	< 39.30	<3.93e1	n/a	89.20	39.20	n/a	n/a
S961003590	A	Manganese - ICP-Acid Digest	ug/g	90.00	<1.00e-02	12.80	13.10	12.95	2.32	3.920	n/a	n/a
S961003590	A	Molybdenum - ICP-Acid Digest	ug/g	94.60	<5.00e-02	56.80	57.60	57.20	1.40	19.60	n/a	n/a
S961003590	A	Sodium - ICP-Acid Digest	ug/g	150.6	1.900	1.88e+05	1.92e+05	2.11	430.0	39.20	n/a	n/a
S961003590	A	Neodymium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	< 39.30	<3.93e1	n/a	97.80	39.20	n/a	n/a
S961003590	A	Nickel - ICP-Acid Digest	ug/g	93.40	<2.00e-02	36.50	43.70	40.10	18.0	7.860	n/a	n/a
S961003590	A	Phosphorus - ICP-Acid Digest	ug/g	89.00	<1.00e-01	6.42e+02	684.0	663.0	103.8	78.60	n/a	n/a
S961003590	A	Lead - ICP-Acid Digest	ug/g	93.00	1.19e-01	3.10e+03	3.15e+03	3.12e+03	1.60	39.20	n/a	n/a
S961003590	A	Sulfur - ICP-Acid Digest	ug/g	88.40	<6.00e-02	< 23.60	<2.36e1	n/a	91.80	23.60	n/a	n/a
S961003590	A	Antimony - ICP-Acid Digest	ug/g	87.40	<1.00e-01	< 39.30	<3.93e1	n/a	93.00	39.20	n/a	n/a
S961003590	A	Selenium - ICP-Acid Digest	ug/g	244.0	5.31e-01	91.20	116.0	103.6	23.9	19.60	n/a	n/a
S961003590	A	Samarium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.930	<3.93e0	n/a	95.20	39.20	n/a	n/a
S961003590	A	Strontium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.930	<3.93e0	n/a	94.00	3.920	n/a	n/a
S961003590	A	Titanium - ICP-Acid Digest	ug/g	88.00	<2.00e-01	< 78.60	<7.86e1	n/a	90.60	78.60	n/a	n/a
S961003590	A	Thallium - ICP-Acid Digest	ug/g	90.70	<5.00e-01	<1.96e+02	<1.96e2	n/a	98.70	196.0	n/a	n/a
S961003590	A	Uranium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	< 19.60	<1.96e1	n/a	94.20	19.60	n/a	n/a
S961003590	A	Zinc - ICP-Acid Digest	ug/g	88.80	1.60e-02	37.20	41.90	39.55	11.9	93.40	n/a	n/a
S961003590	A	Zirconium - ICP-Acid Digest	ug/g	93.20	<1.00e-02	6.640	7.500	7.070	12.2	95.80	3.920	n/a
S961003590	A	Acetate by IC-PIONEX 4000/4500	ug/g	104.8	<2.00e-02	6.68e+02	571.0	619.4	15.7	113.7	83.75	n/a
S961003590	A	Formate by IC-PIONEX 4000/4500	ug/g	101.7	<2.00e-02	1.21e+03	1.11e+03	8.62	103.3	87.93	n/a	n/a
S961003590	A	Fluoride-IC-PIONEX 4000/4500	ug/g	94.98	<1.00e-02	1.78e+03	1.43e+03	21.8	77.80	106.3	n/a	n/a
S961003590	A	Chloride-IC-PIONEX 4000/4500	ug/g	95.32	<1.00e-02	5.89e+03	5.80e+03	1.54	96.71	138.9	n/a	n/a
S961003590	A	Nitrite-IC - PIONEX 4000/4500	ug/g	96.31	<1.07e-01	7.24e+04	7.29e+04	0.69	102.8	874.5	n/a	n/a
S961003590	A	Bromide by Ion Chromatograph	ug/g	94.74	<1.26e-01	<1.03e+03	<1.03e3	n/a	97.45	1.03e+03	n/a	n/a
S961003590	A	Nitrate by IC-PIONEX 4000/4500	ug/g	95.12	1.81e-01	9.54e+04	9.69e+04	1.56	101.3	1.14e+03	n/a	n/a
S961003590	A	Phosphate-IC-PIONEX 4000/4500	ug/g	94.50	<1.19e-01	1.91e+03	1.76e+03	8.17	93.03	971.7	n/a	n/a
S961003590	A	Sulfate by IC-PIONEX 4000/4500	ug/g	96.36	<1.36e-01	8.55e+03	8.74e+03	2.20	95.25	1.11e+03	n/a	n/a
S961003590	A	Oxalate by IC-PIONEX 4000/4500	ug/g	98.29	<1.05e-01	4.64e+03	4.86e+03	4.63	100.0	858.1	n/a	n/a

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk %	Spec %	Det Limit	Count Err%
S94T003707		Bulk Density of Sample	g/mL	n/a	n/a	1.660	n/a	n/a	n/a	n/a	n/a	5.00e-01	n/a
S94T003794		DSC Exotherm using Mettler	Joules/g	104.4	n/a	26.50	24.30	25.40	8.66	n/a	n/a	n/a	n/a
S94T003794		DSC Exotherm DRY Calculated	Joules/g DRY	104.4	n/a	46.90	43.00	44.95	8.68	n/a	n/a	n/a	n/a
S94T003794		% Water by IGA using Mettler	%	100.9	n/a	42.46	44.41	43.44	4.59	n/a	n/a	n/a	n/a
S94T003896		F Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.07e+02	<2.01e2	2.24e+04	n/a	88.30	97.70	207.0	n/a
S94T003896		F Aluminum - ICP-Fusion	ug/g	98.80	<5.00e-02	<2.21e+04	<2.01e3	2.24e+04	2.68	97.70	97.70	1.03e+03	n/a
S94T003896		F Boron - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	94.30	94.30	2.07e+03	n/a
S94T003896		F Arsenic - ICP-Fusion	ug/g	100.4	<1.00e-01	<1.03e+03	<1.00e3	n/a	n/a	94.30	94.30	1.03e+03	n/a
S94T003896		F Barium - ICP-Fusion	ug/g	100.8	<5.00e-03	<1.03e+03	<1.00e3	n/a	n/a	93.90	93.90	1.03e+03	n/a
S94T003896		F Beryllium - ICP-Fusion	ug/g	103.8	<5.00e-03	<1.03e+02	<1.00e2	n/a	n/a	97.70	97.70	103.0	n/a
S94T003896		F Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	98.90	98.90	2.07e+03	n/a
S94T003896		F Cadmium - ICP-Fusion	ug/g	100.4	<1.00e-03	<1.03e+03	<2.01e3	n/a	n/a	97.70	97.70	103.0	n/a
S94T003896		F Cobalt - ICP-Fusion	ug/g	101.4	<2.00e-02	<2.07e+03	<2.01e3	n/a	n/a	97.70	97.70	103.0	n/a
S94T003896		F Cerium - ICP-Fusion	ug/g	100.8	<1.00e-01	<1.03e+03	<2.01e3	n/a	n/a	97.70	97.70	103.0	n/a
S94T003896		F Chromium - ICP-Fusion	ug/g	101.4	<2.00e-02	<2.07e+03	<2.01e3	n/a	n/a	97.70	97.70	103.0	n/a
S94T003896		F Copper - ICP-Fusion	ug/g	104.2	<1.00e-02	<2.07e+03	<2.01e3	2.43e+03	0.82	98.50	97.40	414.0	n/a
S94T003896		F Iron - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.03e+03	<1.00e3	n/a	n/a	97.90	97.90	207.0	n/a
S94T003896		F Potassium - ICP-Fusion	ug/g	100.0	<5.00e-01	7.35e+06	n/a	n/a	n/a	98.00	98.00	1.03e+03	n/a
S94T003896		F Lanthanum - ICP-Fusion	ug/g	102.4	<5.00e-02	<1.03e+03	<1.00e3	n/a	n/a	97.20	97.20	1.03e+03	n/a
S94T003896		F Lithium - ICP-Fusion	ug/g	102.8	<1.00e-02	<2.07e+03	<2.01e3	n/a	n/a	95.90	95.90	207.0	n/a
S94T003896		F Magnesium - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	93.00	93.00	2.07e+03	n/a
S94T003896		F Manganese - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.07e+03	<2.01e3	n/a	n/a	93.20	93.20	207.0	n/a
S94T003896		F Molybdenum - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.03e+03	<1.00e3	n/a	n/a	97.20	97.20	1.03e+03	n/a
S94T003896		F Sodium - ICP-Fusion	ug/g	99.00	1.180	2.41e+05	2.38e+05	2.40e+05	1.25	133.0	133.0	2.07e+03	n/a
S94T003896		F Neodymium - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	97.50	97.50	2.07e+03	n/a
S94T003896		F Nickel - ICP-Fusion	ug/g	101.4	8.50e-02	1.74e+03	2.22e+03	1.98e+03	24.2	97.90	97.90	414.0	n/a
S94T003896		F Phosphorus - ICP-Fusion	ug/g	102.0	<2.00e-01	<4.13e+03	<4.02e3	n/a	n/a	104.0	104.0	4.14e+03	n/a
S94T003896		F Lead - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	97.30	97.30	2.07e+03	n/a
S94T003896		F Sulfur - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.24e+03	<1.20e3	n/a	n/a	92.50	92.50	1.24e+03	n/a
S94T003896		F Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	96.70	96.70	2.07e+03	n/a
S94T003896		F Selenium - ICP-Fusion	ug/g	99.00	<5.00e-02	<1.03e+03	<1.00e3	n/a	n/a	94.80	94.80	1.03e+03	n/a
S94T003896		F Silicon - ICP-Fusion	ug/g	95.40	<5.00e-01	<1.03e+03	<1.00e3	n/a	n/a	93.10	93.10	2.07e+03	n/a
S94T003896		F Samarium - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.07e+03	<2.01e3	n/a	n/a	94.30	94.30	207.0	n/a
S94T003896		F Strontium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.07e+03	<2.01e3	n/a	n/a	90.00	90.00	4.14e+03	n/a
S94T003896		F Tantalum - ICP-Fusion	ug/g	97.00	<5.00e-01	<4.13e+03	<4.02e3	n/a	n/a	87.00	87.00	1.03e+03	n/a
S94T003896		F Thallium - ICP-Fusion	ug/g	96.70	<5.00e-01	<1.03e+04	<1.00e4	n/a	n/a	96.40	96.40	1.03e+03	n/a
S94T003896		F Vanadium - ICP-Fusion	ug/g	100.6	<5.00e-02	<1.03e+03	<2.01e3	n/a	n/a	100.0	100.0	207.0	n/a
S94T003896		F Zinc - ICP-Fusion	ug/g	101.4	<1.00e-02	<2.07e+03	<2.01e3	n/a	n/a	94.10	94.10	207.0	n/a
S94T003896		F Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.07e+03	<2.01e3	15.85	3.15	n/a	n/a	3.880	n/a
S94T003901		A Silver - ICP-Acid Digest	ug/g	23.40	<1.00e-02	16.10	15.60	15.85	3.15	n/a	n/a	3.880	n/a
S94T003901		A Aluminum - ICP-Acid Digest	ug/g	101.4	2.28e-01	2.31e+04	2.19e+04	2.25e+04	5.33	n/a	n/a	19.40	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003901	A	Arsenic - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 38.90	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Boron - ICP-Acid Digest	ug/g	118.2	6.30e-01	72.50	83.80	78.15	14.5	n/a	19.40	n/a
S96T003901	A	Barium - ICP-Acid Digest	ug/g	92.80	<5.00e-02	< 19.40	<1.98e1	n/a	n/a	n/a	19.40	n/a
S96T003901	A	Bismuth - ICP-Acid Digest	ug/g	94.60	<1.00e-03	< 1.940	<1.98e0	n/a	n/a	n/a	1.940	n/a
S96T003901	A	Calcium - ICP-Acid Digest	ug/g	92.40	<5.00e-01	< 38.90	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Calcium - ICP-Acid Digest	ug/g	101.8	3.80e-01	4.20e+02	328.0	374.0	24.6	n/a	38.80	n/a
S96T003901	A	Cadmium - ICP-Acid Digest	ug/g	90.80	<5.00e-03	20.90	19.40	20.15	7.44	n/a	1.940	n/a
S96T003901	A	Cerium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 38.90	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Cobalt - ICP-Acid Digest	ug/g	93.80	<2.00e-02	< 7.780	<7.92e0	n/a	n/a	n/a	7.780	n/a
S96T003901	A	Chromium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	2.51e+03	2.38e+03	2.44e+03	5.32	n/a	3.880	n/a
S96T003901	A	Copper - ICP-Acid Digest	ug/g	92.00	1.30e-02	6.000	6.170	6.085	2.79	n/a	3.880	n/a
S96T003901	A	Iron - ICP-Acid Digest	ug/g	93.00	<5.00e-02	1.79e+02	171.0	175.0	4.57	n/a	19.40	n/a
S96T003901	A	Potassium - ICP-Acid Digest	ug/g	93.80	<5.00e-01	3.61e+03	3.42e+03	3.52e+03	5.41	n/a	19.40	n/a
S96T003901	A	Lanthanum - ICP-Acid Digest	ug/g	95.00	<1.00e-02	< 19.40	<1.98e1	n/a	n/a	n/a	19.40	n/a
S96T003901	A	Lithium - ICP-Acid Digest	ug/g	96.60	<5.00e-02	7.250	4.050	5.650	56.6	n/a	3.880	n/a
S96T003901	A	Magnesium - ICP-Acid Digest	ug/g	88.60	<1.00e-01	< 38.90	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Manganese - ICP-Acid Digest	ug/g	90.00	<1.00e-02	30.00	28.60	29.30	4.78	n/a	38.80	n/a
S96T003901	A	Molybdenum - ICP-Acid Digest	ug/g	94.60	<5.00e-02	54.30	50.60	52.45	7.05	n/a	19.40	n/a
S96T003901	A	Sodium - ICP-Acid Digest	ug/g	150.6	1.900	1.98e+05	1.95e+05	1.96e+05	1.53	n/a	38.80	n/a
S96T003901	A	Neodymium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	< 38.90	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Nickel - ICP-Acid Digest	ug/g	93.40	<2.00e-02	72.50	71.10	71.80	1.95	n/a	7.780	n/a
S96T003901	A	Phosphorus - ICP-Acid Digest	ug/g	93.40	<2.00e-01	7.89e+02	762.0	775.5	3.48	n/a	77.80	n/a
S96T003901	A	Lead - ICP-Acid Digest	ug/g	89.00	<1.00e-01	47.30	<3.96e1	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Sulfur - ICP-Acid Digest	ug/g	93.00	1.19e-01	3.30e+03	3.23e+03	3.26e+03	2.14	n/a	38.80	n/a
S96T003901	A	Selenium - ICP-Acid Digest	ug/g	88.40	<6.00e-02	< 23.30	<2.38e1	n/a	n/a	n/a	23.40	n/a
S96T003901	A	Silicon - ICP-Acid Digest	ug/g	244.0	5.31e-01	1.03e+02	145.0	124.0	33.9	n/a	38.80	n/a
S96T003901	A	Samarium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 38.90	<3.96e1	n/a	n/a	n/a	19.40	n/a
S96T003901	A	Strontium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.890	<3.96e0	n/a	n/a	n/a	38.80	n/a
S96T003901	A	Titanium - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 3.890	<3.96e0	n/a	n/a	n/a	3.880	n/a
S96T003901	A	Thallium - ICP-Acid Digest	ug/g	88.00	<2.00e-01	< 77.80	<7.92e1	n/a	n/a	n/a	77.80	n/a
S96T003901	A	Uranium - ICP-Acid Digest	ug/g	90.70	<5.00e-01	<1.94e+02	<1.98e2	n/a	n/a	n/a	194.0	n/a
S96T003901	A	Vanadium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	< 19.40	<1.98e1	n/a	n/a	n/a	19.40	n/a
S96T003901	A	Zinc - ICP-Acid Digest	ug/g	88.80	1.60e-02	40.00	38.90	39.45	2.79	n/a	3.880	n/a
S96T003901	A	Zirconium - ICP-Acid Digest	ug/g	93.20	<1.00e-02	8.920	8.770	8.845	1.70	n/a	3.880	n/a
S96T003910	W	Acetate by IC-DIONEX 4000/4500	ug/g	104.8	<2.00e-02	6.23e+02	599.0	610.8	3.93	n/a	85.51	n/a
S96T003910	W	Formate by IC-DIONEX 4000/4500	ug/g	101.7	<2.00e-02	1.18e+03	1.12e+03	1.15e+03	5.22	n/a	89.78	n/a
S96T003910	W	Fluoride-IC-DIONEX 4000/4500	ug/g	94.58	<1.30e-02	1.84e+03	1.84e+03	1.84e+03	0.00	n/a	108.5	n/a
S96T003910	W	Chloride-IC-DIONEX 4000/4500	ug/g	95.32	<1.70e-02	5.32e+03	5.13e+03	5.22e+03	3.64	n/a	141.9	n/a
S96T003910	W	Nitrite-IC - Dionex 4000/4500	ug/g	96.31	<1.07e-01	6.62e+04	6.59e+04	6.59e+04	0.91	n/a	893.0	n/a
S96T003910	W	Bromide by Ion Chromatograph	ug/g	94.74	<1.26e-01	<1.05e+03	<1.02e3	n/a	n/a	n/a	1.05e+03	n/a
S96T003910	W	Nitrate by IC-DIONEX 4000/4500	ug/g	95.12	1.81e-01	1.12e+05	1.12e+05	1.12e+05	0.00	n/a	1.17e+03	n/a
S96T003910	W	Phosphate-IC-DIONEX 4000/4500	ug/g	94.50	<1.19e-01	2.26e+03	2.69e+03	2.48e+03	17.4	n/a	992.6	n/a
S96T003910	W	Sulfate by IC-DIONEX 4000/4500	ug/g	96.36	<1.36e-01	8.99e+03	8.72e+03	8.86e+03	3.05	n/a	1.13e+03	n/a
S96T003910	W	Oxalate by IC-DIONEX 4000/4500	ug/g	98.29	<1.05e-01	5.00e+03	5.01e+03	5.01e+03	0.20	n/a	876.6	n/a

L Lower Half of Segment: L Lower Half of Segment

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003708		Bulk Density of Sample	g/mL	n/a	n/a	1.640	n/a	n/a	n/a	n/a	5.00e-01	n/a
S96T003795		DSC Exotherm using Mettler	Joules/g	104.4	n/a	30.70	32.70	31.70	6.31	n/a	n/a	n/a
S96T003795		DSC Exotherm Dry Calculated	Joules/g	n/a	n/a	55.40	59.00	57.20	6.29	n/a	n/a	n/a
S96T003795		% Water by TGA using Mettler	%	100.0	n/a	43.23	45.83	44.53	5.84	n/a	n/a	n/a
S96T003795		F Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.06e-02	<2.07e2	0.84	n/a	86.20	206.0	n/a
S96T003890		F Aluminum - ICP-Fusion	ug/g	98.80	<5.00e-01	2.38e+04	2.40e+04	2.39e+04	0.84	91.00	1.03e+03	n/a
S96T003890		F Arsenic - ICP-Fusion	ug/g	104.4	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	100.0	2.06e+03	n/a
S96T003890		F Boron - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	94.60	1.03e+03	n/a
S96T003890		F Barium - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	95.60	1.03e+03	n/a
S96T003890		F Beryllium - ICP-Fusion	ug/g	103.8	<1.00e-01	<1.03e+02	<1.04e2	n/a	n/a	97.10	103.0	n/a
S96T003890		F Bismuth - ICP-Fusion	ug/g	104.4	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	103.0	2.06e+03	n/a
S96T003890		F Calcium - ICP-Fusion	ug/g	100.4	<5.00e-03	<1.03e+02	<1.04e2	n/a	n/a	98.10	2.06e+03	n/a
S96T003890		F Cadmium - ICP-Fusion	ug/g	101.4	<5.00e-03	<1.03e+02	<1.04e2	n/a	n/a	97.00	103.0	n/a
S96T003890		F Cerium - ICP-Fusion	ug/g	101.4	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	96.30	2.06e+03	n/a
S96T003890		F Cobalt - ICP-Fusion	ug/g	101.8	<1.00e-02	<4.12e+02	<4.14e2	n/a	n/a	96.30	410.0	n/a
S96T003890		F Chromium - ICP-Fusion	ug/g	101.8	<1.00e-02	<2.26e+03	<2.26e+03	2.29e+03	2.62	97.20	206.0	n/a
S96T003890		F Copper - ICP-Fusion	ug/g	104.2	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	97.50	206.0	n/a
S96T003890		F Iron - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	96.60	1.03e+03	n/a
S96T003890		F Potassium - ICP-Fusion	ug/g	100.0	<5.00e-01	1.04e+07	n/a	n/a	n/a	n/a	1.03e+04	n/a
S96T003890		F Lanthanum - ICP-Fusion	ug/g	102.4	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	96.50	1.03e+03	n/a
S96T003890		F Lithium - ICP-Fusion	ug/g	102.8	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	95.80	206.0	n/a
S96T003890		F Magnesium - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	92.60	2.06e+03	n/a
S96T003890		F Manganese - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	92.30	206.0	n/a
S96T003890		F Molybdenum - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	96.60	1.03e+03	n/a
S96T003890		F Sodium - ICP-Fusion	ug/g	99.00	<1.180	2.58e+05	2.45e+05	2.52e+05	5.17	86.90	1.03e+03	n/a
S96T003890		F Neodymium - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	96.90	2.06e+03	n/a
S96T003890		F Nickel - ICP-Fusion	ug/g	101.4	8.50e-02	2.34e+03	1.46e+03	1.90e+03	46.3	97.70	410.0	n/a
S96T003890		F Phosphorus - ICP-Fusion	ug/g	102.0	<2.00e-01	<4.12e+03	<4.14e3	n/a	n/a	102.0	4.10e+03	n/a
S96T003890		F Lead - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	97.70	2.06e+03	n/a
S96T003890		F Sulfur - ICP-Fusion	ug/g	101.2	<1.00e-01	3.27e+03	3.29e+03	3.28e+03	0.61	95.90	2.06e+03	n/a
S96T003890		F Antimony - ICP-Fusion	ug/g	96.00	<5.00e-02	<1.24e+03	<1.24e3	n/a	n/a	92.90	1.23e+03	n/a
S96T003890		F Selenium - ICP-Fusion	ug/g	97.40	<5.00e-02	<1.03e+03	<2.07e3	n/a	n/a	96.30	2.06e+03	n/a
S96T003890		F Silver - ICP-Fusion	ug/g	95.40	<5.00e-02	<1.03e+03	<1.04e3	n/a	n/a	94.20	1.03e+03	n/a
S96T003890		F Smaerium - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.06e+03	<2.07e3	n/a	n/a	92.40	2.06e+03	n/a
S96T003890		F Strontium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	94.00	206.0	n/a
S96T003890		F Tantalum - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	93.20	206.0	n/a
S96T003890		F Thallium - ICP-Fusion	ug/g	97.00	<2.00e-01	<4.12e+03	<4.14e3	n/a	n/a	90.70	4.10e+03	n/a
S96T003890		F Vanadium - ICP-Fusion	ug/g	96.70	<5.00e-01	<1.03e+04	<1.04e4	n/a	n/a	85.00	1.03e+04	n/a
S96T003890		F Zinc - ICP-Fusion	ug/g	100.6	<5.00e-02	<2.06e+02	<2.07e2	n/a	n/a	95.70	1.03e+03	n/a
S96T003890		F Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	99.00	206.0	n/a
S96T003890		F Alpha of Digested Solid	ug/g	97.66	<7.2e-02	<1.91e-02	<1.92E-2	n/a	n/a	n/a	4.80e-02	5.00E-02
S96T003902		A Silver - ICP-Acid Digest	ug/g	24.80	<1.00e-02	16.00	15.80	15.90	1.26	31.50	4.000	n/a
S96T003902		A Aluminum - ICP-Acid Digest	ug/g	101.2	<1.44e-01	2.31e+04	2.41e+04	2.34e+04	4.24	77.60	40.00	n/a
S96T003902		A Arsenic - ICP-Acid Digest	ug/g	98.00	<1.00e-01	<40.00	<4.04e1	n/a	n/a	98.00	40.00	n/a
S96T003902		A Barium - ICP-Acid Digest	ug/g	121.2	<4.34e-01	1.08e+02	69.60	88.80	43.2	98.80	20.00	n/a
S96T003902		A Bismuth - ICP-Acid Digest	ug/g	98.60	<5.00e-02	<20.00	<2.02e1	n/a	n/a	98.20	20.00	n/a
S96T003902		A Beryllium - ICP-Acid Digest	ug/g	97.40	<5.00e-03	<2.000	<2.02e0	n/a	n/a	96.20	2.000	n/a

Sample#	R/A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003502	A	Bismuth - ICP-Acid Digest	ug/g	95.20	< 1.00e-01	< 40.00	< 4.04e1	N/A	N/A	96.60	40.00	N/A
S961003502	A	Calcium - ICP-Acid Digest	ug/g	100.2	2.320	2.40e+02	237.0	238.5	1.26	102.8	40.00	N/A
S961003502	A	Cadmium - ICP-Acid Digest	ug/g	93.80	< 5.00e-01	12.90	13.00	12.95	0.77	95.80	2.000	N/A
S961003502	A	Cerium - ICP-Acid Digest	ug/g	100.6	< 1.00e-01	< 40.00	< 4.04e1	N/A	N/A	100.6	40.00	N/A
S961003502	A	Cobalt - ICP-Acid Digest	ug/g	96.80	< 1.00e-02	< 7.900	< 8.09e0	N/A	N/A	97.40	8.000	N/A
S961003502	A	Chromium - ICP-Acid Digest	ug/g	95.20	< 1.00e-02	2.25e+03	2.31e+03	2.28e+03	2.63	93.40	4.000	N/A
S961003502	A	Copper - ICP-Acid Digest	ug/g	95.40	< 1.00e-01	5.290	10.50	7.895	66.0	94.40	4.000	N/A
S961003502	A	Iron - ICP-Acid Digest	ug/g	98.40	7.00e-02	78.70	74.80	76.75	5.08	94.60	20.00	N/A
S961003502	A	Potassium - ICP-Acid Digest	ug/g	99.40	< 5.00e-01	< 20.00	< 2.02e1	N/A	N/A	87.00	200.0	N/A
S961003502	A	Lanthanum - ICP-Acid Digest	ug/g	91.00	< 1.00e-02	< 40.00	< 4.04e1	12.40	6.45	101.4	4.000	N/A
S961003502	A	Lithium - ICP-Acid Digest	ug/g	100.4	< 1.00e-02	11.10	11.10	11.10	0.00	92.80	4.000	N/A
S961003502	A	Magnesium - ICP-Acid Digest	ug/g	96.40	< 5.00e-02	53.40	54.50	53.95	2.04	98.20	20.00	N/A
S961003502	A	Molybdenum - ICP-Acid Digest	ug/g	138.6	1.110	1.97e+05	1.95e+05	1.96e+05	1.02	-3.26e+2	40.00	N/A
S961003502	A	Sodium - ICP-Acid Digest	ug/g	99.40	< 1.00e-01	< 40.00	< 4.04e1	N/A	N/A	99.40	40.00	N/A
S961003502	A	Nickel - ICP-Acid Digest	ug/g	96.60	< 2.00e-02	55.70	52.00	53.85	6.87	96.80	8.000	N/A
S961003502	A	Phosphorus - ICP-Acid Digest	ug/g	94.60	< 2.00e-01	8.49e+02	925.0	887.0	8.57	146.6	80.00	N/A
S961003502	A	Lead - ICP-Acid Digest	ug/g	93.20	< 1.00e-01	< 4.00	< 2.43e1	N/A	N/A	93.00	40.00	N/A
S961003502	A	Sulfur - ICP-Acid Digest	ug/g	89.40	< 5.00e-02	< 24.00	< 2.43e1	N/A	N/A	93.60	24.00	N/A
S961003502	A	Antimony - ICP-Acid Digest	ug/g	89.20	< 1.00e-01	< 40.00	< 4.04e1	N/A	N/A	93.60	40.00	N/A
S961003502	A	Selenium - ICP-Acid Digest	ug/g	264.0	2.20e-01	1.41e+02	86.60	113.8	47.8	90.40	20.00	N/A
S961003502	A	Silicon - ICP-Acid Digest	ug/g	97.20	< 1.00e-01	< 40.00	< 4.04e1	N/A	N/A	97.00	40.00	N/A
S961003502	A	Samarium - ICP-Acid Digest	ug/g	97.20	< 1.00e-02	< 4.000	< 4.04e0	N/A	N/A	97.00	4.000	N/A
S961003502	A	Titanium - ICP-Acid Digest	ug/g	95.20	< 1.00e-02	< 4.000	< 4.04e0	N/A	N/A	95.40	4.000	N/A
S961003502	A	Thallium - ICP-Acid Digest	ug/g	89.80	< 2.00e-01	< 79.90	< 8.09e0	N/A	N/A	90.60	80.00	N/A
S961003502	A	Uranium - ICP-Acid Digest	ug/g	95.40	< 5.00e-01	< 2.00e+02	< 2.02e2	N/A	N/A	90.60	200.0	N/A
S961003502	A	Vanadium - ICP-Acid Digest	ug/g	95.00	< 5.00e-02	< 20.00	< 2.02e1	N/A	N/A	93.60	4.000	N/A
S961003502	A	Zinc - ICP-Acid Digest	ug/g	90.20	3.90e-02	34.80	38.00	36.40	8.70	95.60	20.00	N/A
S961003502	A	Zirconium - ICP-Acid Digest	ug/g	96.40	< 2.00e-02	7.800	7.430	7.615	4.86	97.80	4.000	N/A
S961003511	A	Acetate by IC-Dionex 4000/4500	ug/g	110.5	< 2.00e-02	7.85e+02	437.0	610.8	57.0	N/A	3.949	N/A
S961003511	A	Fluoride by IC-Dionex 4000/4500	ug/g	88.93	< 2.10e-02	7.7e+02	888.0	830.4	13.8	N/A	4.146	N/A
S961003511	A	Chloride by IC-Dionex 4000/4500	ug/g	88.98	< 1.00e-02	2.65e+03	2.65e+03	2.65e+03	0.38	N/A	105.2	N/A
S961003511	A	Nitrate by IC-Dionex 4000/4500	ug/g	92.66	< 1.70e-02	5.18e+03	5.08e+03	5.08e+03	3.94	N/A	137.2	N/A
S961003511	A	Bromide by Ion Chromatograph	ug/g	93.54	< 1.07e-01	6.52e+04	6.37e+04	6.45e+04	2.33	N/A	865.3	N/A
S961003511	A	Nitrate by IC-Dionex 4000/4500	ug/g	91.68	< 1.26e-01	< 1.02e+03	< 1.00e3	N/A	N/A	N/A	1.02e+03	N/A
S961003511	A	Phosphate by IC-Dionex 4000/4500	ug/g	92.93	1.75e-01	1.18e+05	1.14e+05	1.16e+05	3.45	N/A	1.15e+03	N/A
S961003511	A	Sulfate by IC-Dionex 4000/4500	ug/g	90.09	< 1.19e-01	1.70e+03	1.42e+03	1.56e+03	17.9	N/A	1.94e3	N/A
S961003511	A	Oxalate by IC-Dionex 4000/4500	ug/g	92.55	< 1.36e-01	7.79e+03	8.03e+03	7.91e+03	3.03	N/A	1.10e+3	N/A
S961003511	A	Oxalate by IC-Dionex 4000/4500	ug/g	98.09	< 1.05e-01	5.21e+03	5.90e+03	5.55e+03	12.4	N/A	849.3	N/A

CORE NUMBER: 152
SEGMENT #: 20

SEGMENT PORTION: U Upper Half of Segment

Table 1. Data Summary Table
AN-105

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003710		Bulk Density of Sample	g/mL	101.6	n/a	1.580	39.70	39.10	n/a	n/a	5.00e-01	n/a
S961003796		DSC Exotherm using Mettler	Joules/g	n/a	n/a	38.50	79.70	39.10	3.07	n/a	n/a	n/a
S961003796		DSC Exotherm DRY, Calculated	Joules/g Dry	n/a	n/a	70.00	73.00	71.50	3.06	n/a	n/a	n/a
S961003796		% Water by TGA using Mettler	%	100.2	n/a	45.15	46.04	45.59	1.95	n/a	n/a	n/a
S961003897	F	Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.12e+02	<2.07e2	2.34e+04	1.71	85.10	212.0	n/a
S961003897	F	Aluminum - ICP-Fusion	ug/g	98.80	<5.00e-02	2.32e+04	2.32e+04	2.34e+04	1.71	92.60	1.06e+03	n/a
S961003897	F	Arsenic - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	101.0	1.12e+03	n/a
S961003897	F	Boron - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	95.80	2.12e+03	n/a
S961003897	F	Barium - ICP-Fusion	ug/g	100.8	<5.00e-03	<1.06e+03	<1.04e3	n/a	n/a	94.40	1.06e+03	n/a
S961003897	F	Beryllium - ICP-Fusion	ug/g	103.8	<5.00e-03	<1.06e+02	<1.04e2	n/a	n/a	99.80	2.12e+03	n/a
S961003897	F	Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	97.70	2.12e+03	n/a
S961003897	F	Calcium - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	97.70	106.0	n/a
S961003897	F	Cadmium - ICP-Fusion	ug/g	101.4	<5.00e-03	<1.06e+02	<1.04e2	n/a	n/a	96.20	2.12e+03	n/a
S961003897	F	Cerium - ICP-Fusion	ug/g	100.8	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	97.60	422.0	n/a
S961003897	F	Cobalt - ICP-Fusion	ug/g	101.4	<2.00e-02	<4.24e+02	<4.14e2	3.88e+03	3.61	98.30	212.0	n/a
S961003897	F	Chromium - ICP-Fusion	ug/g	104.2	<1.00e-02	3.95e+03	3.81e+03	3.88e+03	3.61	98.60	212.0	n/a
S961003897	F	Copper - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	97.70	1.06e+03	n/a
S961003897	F	Iron - ICP-Fusion	ug/g	101.6	<5.00e-01	5.81e+06	n/a	n/a	n/a	n/a	1.06e+04	n/a
S961003897	F	Potassium - ICP-Fusion	ug/g	102.4	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	97.30	1.06e+03	n/a
S961003897	F	Lanthanum - ICP-Fusion	ug/g	102.4	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	96.80	212.0	n/a
S961003897	F	Lithium - ICP-Fusion	ug/g	102.8	<1.00e-02	<2.12e+02	<2.07e2	n/a	n/a	94.50	2.12e+03	n/a
S961003897	F	Magnesium - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	93.20	212.0	n/a
S961003897	F	Manganese - ICP-Fusion	ug/g	99.0	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	97.50	1.06e+03	n/a
S961003897	F	Molybdenum - ICP-Fusion	ug/g	98.0	<1.00e-02	<1.06e+03	<1.04e3	2.28e+05	0.44	84.80	2.12e+03	n/a
S961003897	F	Sodium - ICP-Fusion	ug/g	99.00	1.180	2.28e+05	2.27e+05	2.28e+05	0.44	84.80	2.12e+03	n/a
S961003897	F	Neodymium - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	97.70	422.0	n/a
S961003897	F	Nickel - ICP-Fusion	ug/g	101.4	8.50e-02	1.33e+03	1.54e+03	1.44e+03	14.6	97.60	422.0	n/a
S961003897	F	Phosphorus - ICP-Fusion	ug/g	102.0	<2.00e-01	<4.24e+03	<4.14e3	n/a	n/a	106.0	4.22e+03	n/a
S961003897	F	Lead - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.12e+03	<2.07e3	3.06e+03	2.28	96.40	2.12e+03	n/a
S961003897	F	Sulfur - ICP-Fusion	ug/g	96.00	<5.00e-02	<2.12e+03	<1.24e3	n/a	n/a	98.10	2.12e+03	n/a
S961003897	F	Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	98.10	2.12e+03	n/a
S961003897	F	Selenium - ICP-Fusion	ug/g	95.40	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	94.20	1.06e+03	n/a
S961003897	F	Silicon - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.12e+03	<2.07e3	n/a	n/a	93.40	2.12e+03	n/a
S961003897	F	Strontium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.12e+02	<2.07e2	n/a	n/a	94.70	212.0	n/a
S961003897	F	Titanium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.12e+02	<2.07e2	n/a	n/a	94.40	212.0	n/a
S961003897	F	Thallium - ICP-Fusion	ug/g	97.00	<5.00e-01	<4.24e+03	<4.14e3	n/a	n/a	92.60	4.22e+03	n/a
S961003897	F	Uranium - ICP-Fusion	ug/g	96.70	<5.00e-01	<1.06e+04	<1.04e4	n/a	n/a	88.00	1.06e+04	n/a
S961003897	F	Vanadium - ICP-Fusion	ug/g	100.6	<5.00e-02	<1.06e+03	<1.04e3	n/a	n/a	96.60	1.06e+03	n/a
S961003897	F	Zinc - ICP-Fusion	ug/g	101.4	<1.00e-02	<2.12e+02	<2.07e2	n/a	n/a	98.80	212.0	n/a
S961003897	F	Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.12e+02	<2.07e2	n/a	n/a	94.30	212.0	n/a
S961003903	A	Silver - ICP-Acid Digest	ug/g	24.80	<1.00e-02	15.10	14.90	15.00	1.33	n/a	3.820	n/a
S961003903	A	Aluminum - ICP-Acid Digest	ug/g	101.2	1.44e-01	2.48e+04	2.44e+04	2.46e+04	1.63	n/a	19.10	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Average	RPP %	Spk Rec %	Det Limit	Count	Err%
5961003903	A	Arsenic - ICP-Acid Digest	ug/g	98.00	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Boron - ICP-Acid Digest	ug/g	121.2	4.34e-01	89.60	77.30	83.45	14.7	19.10	n/a	n/a
5961003903	A	Barium - ICP-Acid Digest	ug/g	98.60	<5.00e-02	< 19.10	<1.88e1	n/a	n/a	19.10	n/a	n/a
5961003903	A	Beryllium - ICP-Acid Digest	ug/g	97.40	<1.00e-03	< 1.910	<1.88e0	n/a	n/a	1.910	n/a	n/a
5961003903	A	Bismuth - ICP-Acid Digest	ug/g	95.20	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Calcium - ICP-Acid Digest	ug/g	100.2	2.320	2.53e+02	288.0	255.5	1.96	38.20	n/a	n/a
5961003903	A	Cadmium - ICP-Acid Digest	ug/g	93.80	<5.00e-03	< 13.30	13.20	13.40	2.99	1.910	n/a	n/a
5961003903	A	Cerium - ICP-Acid Digest	ug/g	100.6	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Cobalt - ICP-Acid Digest	ug/g	96.80	<2.00e-02	< 7.660	<7.52e0	n/a	n/a	7.660	n/a	n/a
5961003903	A	Chromium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	3.74e+03	3.38e+03	3.42e+03	2.63	3.820	n/a	n/a
5961003903	A	Copper - ICP-Acid Digest	ug/g	95.40	<1.00e-02	6.900	5.910	6.405	15.5	3.820	n/a	n/a
5961003903	A	Iron - ICP-Acid Digest	ug/g	96.40	7.70e-02	82.10	90.70	86.40	9.95	19.10	n/a	n/a
5961003903	A	Potassium - ICP-Acid Digest	ug/g	98.40	<5.00e-01	3.73e+03	3.81e+03	3.89e+03	4.20	19.10	n/a	n/a
5961003903	A	Lanthanum - ICP-Acid Digest	ug/g	100.4	<5.00e-02	< 19.10	<1.88e1	n/a	n/a	19.10	n/a	n/a
5961003903	A	Lithium - ICP-Acid Digest	ug/g	99.40	<1.00e-02	< 12.00	12.00	12.10	1.65	3.820	n/a	n/a
5961003903	A	Magnesium - ICP-Acid Digest	ug/g	91.00	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Manganese - ICP-Acid Digest	ug/g	93.00	<1.00e-02	9.470	9.450	9.460	0.21	3.820	n/a	n/a
5961003903	A	Molybdenum - ICP-Acid Digest	ug/g	96.40	<5.00e-02	55.60	57.40	56.50	3.19	19.10	n/a	n/a
5961003903	A	Sodium - ICP-Acid Digest	ug/g	138.6	1.110	1.91e+05	1.86e+05	1.88e+05	2.65	38.20	n/a	n/a
5961003903	A	Nickel - ICP-Acid Digest	ug/g	99.40	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Neodymium - ICP-Acid Digest	ug/g	96.60	<2.00e-02	59.10	54.30	56.70	8.47	7.660	n/a	n/a
5961003903	A	Phosphorus - ICP-Acid Digest	ug/g	94.60	<2.00e-01	6.94e+02	578.0	636.0	18.2	76.60	n/a	n/a
5961003903	A	Lead - ICP-Acid Digest	ug/g	90.80	<1.00e-01	43.20	42.90	43.05	0.70	38.20	n/a	n/a
5961003903	A	Sulfur - ICP-Acid Digest	ug/g	93.20	<1.00e-01	2.74e+03	2.72e+03	2.74e+03	0.73	38.20	n/a	n/a
5961003903	A	Antimony - ICP-Acid Digest	ug/g	89.40	<6.00e-02	< 23.00	<2.28e1	n/a	n/a	23.00	n/a	n/a
5961003903	A	Selenium - ICP-Acid Digest	ug/g	89.20	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Silicon - ICP-Acid Digest	ug/g	264.0	2.20e-01	1.49e+02	147.0	148.0	1.35	19.10	n/a	n/a
5961003903	A	Samarium - ICP-Acid Digest	ug/g	97.20	<1.00e-01	< 38.30	<3.76e1	n/a	n/a	38.20	n/a	n/a
5961003903	A	Strontium - ICP-Acid Digest	ug/g	97.20	<1.00e-02	< 3.830	<3.76e0	n/a	n/a	3.820	n/a	n/a
5961003903	A	Titanium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	< 3.830	<3.76e0	n/a	n/a	3.820	n/a	n/a
5961003903	A	Thallium - ICP-Acid Digest	ug/g	89.80	<2.00e-01	< 76.60	<7.52e1	n/a	n/a	76.60	n/a	n/a
5961003903	A	Uranium - ICP-Acid Digest	ug/g	95.40	<5.00e-01	<1.91e+02	<1.88e2	n/a	n/a	19.10	n/a	n/a
5961003903	A	Vanadium - ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 19.10	<1.88e1	n/a	n/a	19.10	n/a	n/a
5961003903	A	Zinc - ICP-Acid Digest	ug/g	90.20	<3.90e-02	37.00	38.30	37.65	3.45	3.820	n/a	n/a
5961003903	A	Zirconium - ICP-Acid Digest	ug/g	96.40	<1.00e-02	7.500	7.560	7.530	0.80	3.820	n/a	n/a
5961003912	W	Acetate by IC-BIOMEX 4000/4500	ug/g	110.5	<2.00e-02	7.98e+02	450.0	928.6	8.51	4.047	n/a	n/a
5961003912	W	Formate by IC-BIOMEX 4000/4500	ug/g	98.33	<2.00e-02	8.89e+02	968.0	928.6	8.51	4.269	n/a	n/a
5961003912	W	Fluoride-IC-BIOMEX 4000/4500	ug/g	88.98	<1.30e-02	2.80e+03	2.51e+03	2.66e+03	10.9	107.9	n/a	n/a
5961003912	W	Chloride-IC-BIOMEX 4000/4500	ug/g	92.66	<1.70e-02	5.24e+03	5.18e+03	5.21e+03	1.15	141.0	n/a	n/a
5961003912	W	Nitrate-IC - BIOMEX 4000/4500	ug/g	93.34	<1.07e-01	6.49e+04	6.44e+04	6.46e+04	0.77	887.6	n/a	n/a
5961003912	W	Bromide by Ion Chromatograph	ug/g	91.68	<1.26e-01	<1.04e+03	<1.07e3	n/a	n/a	1.05e+03	n/a	n/a
5961003912	W	Nitrate by IC-BIOMEX 4000/4500	ug/g	92.95	<1.79e-01	9.07e+04	9.92e+04	9.49e+04	8.95	1.16e+03	n/a	n/a
5961003912	W	Phosphate-IC-BIOMEX 4000/4500	ug/g	90.09	<1.19e-01	2.61e+03	2.61e+03	2.66e+03	31.0	986.5	n/a	n/a
5961003912	W	Sulfate by IC-BIOMEX 4000/4500	ug/g	92.55	<1.36e-01	8.34e+03	7.88e+03	7.88e+03	12.6	1.13e+03	n/a	n/a
5961003912	W	Oxalate by IC-BIOMEX 4000/4500	ug/g	98.09	<1.09e-01	1.01e+04	1.01e+04	1.07e+04	0.00	871.2	n/a	n/a

L Lower Half of Segment: L Lower Half of Segment

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S96T003711		Bulk Density of Sample	g/mL	n/a	n/a	1.600	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S96T003797		Bulk Density of Sample	Joules/g	101.6	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S96T003797		DSC Exotherm Dry Calculated	Joules/g	101.6	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S96T003797		% Water by TGA using Mettler	%	100.2	n/a	36.14	39.33	37.73	8.45	n/a	204.0	n/a	n/a
S96T003891		F Silver - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.04e+02	<2.06e2	1.88e+04	3.73	91.40	1.02e+03	n/a	n/a
S96T003891		F Aluminum - ICP-Fusion	ug/g	96.20	<5.00e-02	1.91e+04	1.84e+04	1.88e+04	3.73	91.40	1.02e+03	n/a	n/a
S96T003891		F Arsenic - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	98.80	2.04e+03	n/a	n/a
S96T003891		F Boron - ICP-Fusion	ug/g	98.00	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	95.50	1.02e+03	n/a	n/a
S96T003891		F Barium - ICP-Fusion	ug/g	98.20	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	95.50	1.02e+03	n/a	n/a
S96T003891		F Beryllium - ICP-Fusion	ug/g	100.4	<1.00e-03	<1.02e+02	<1.03e2	n/a	n/a	96.00	102.0	n/a	n/a
S96T003891		F Bismuth - ICP-Fusion	ug/g	101.4	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	97.10	2.04e+03	n/a	n/a
S96T003891		F Calcium - ICP-Fusion	ug/g	100.8	<1.00e-01	<1.02e+02	<1.03e2	n/a	n/a	102.0	2.04e+03	n/a	n/a
S96T003891		F Cadmium - ICP-Fusion	ug/g	100.2	<5.00e-03	<1.02e+02	<1.03e2	n/a	n/a	97.90	102.0	n/a	n/a
S96T003891		F Cerium - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	95.90	2.04e+03	n/a	n/a
S96T003891		F Cobalt - ICP-Fusion	ug/g	100.6	<1.00e-02	<4.08e+02	<4.11e2	1.94e+03	3.60	98.10	204.0	n/a	n/a
S96T003891		F Chromium - ICP-Fusion	ug/g	100.8	<1.00e-02	1.98e+03	1.91e+03	1.94e+03	3.60	98.10	204.0	n/a	n/a
S96T003891		F Copper - ICP-Fusion	ug/g	101.0	<1.00e-02	<2.04e+02	<2.06e2	n/a	n/a	96.50	204.0	n/a	n/a
S96T003891		F Iron - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	97.50	1.02e+03	n/a	n/a
S96T003891		F Potassium - ICP-Fusion	ug/g	94.60	<5.00e-01	7.24e+06	n/a	n/a	n/a	97.50	1.02e+04	n/a	n/a
S96T003891		F Lanthanum - ICP-Fusion	ug/g	100.2	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	96.80	1.02e+03	n/a	n/a
S96T003891		F Lithium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.04e+02	<2.06e2	n/a	n/a	94.20	204.0	n/a	n/a
S96T003891		F Magnesium - ICP-Fusion	ug/g	95.60	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	92.20	2.04e+03	n/a	n/a
S96T003891		F Manganese - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.04e+02	<2.06e2	n/a	n/a	95.40	204.0	n/a	n/a
S96T003891		F Molybdenum - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	97.30	1.02e+03	n/a	n/a
S96T003891		F Sodium - ICP-Fusion	ug/g	96.20	1.320	2.32e+05	2.25e+05	2.28e+05	3.06	81.70	2.04e+03	n/a	n/a
S96T003891		F Neodymium - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	97.00	2.04e+03	n/a	n/a
S96T003891		F Nickel - ICP-Fusion	ug/g	99.60	2.38e-01	1.50e+03	857.0	1.23e+03	60.5	96.80	408.0	n/a	n/a
S96T003891		F Phosphorus - ICP-Fusion	ug/g	101.4	<2.00e-01	<4.08e+03	<4.11e3	n/a	n/a	109.0	4.08e+03	n/a	n/a
S96T003891		F Lead - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	98.90	2.04e+03	n/a	n/a
S96T003891		F Sulfur - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.28e+03	<2.49e+03	2.48e+03	0.40	96.70	2.04e+03	n/a	n/a
S96T003891		F Antimony - ICP-Fusion	ug/g	94.60	<5.00e-02	<1.02e+03	<1.23e3	n/a	n/a	92.10	1.23e+03	n/a	n/a
S96T003891		F Selenium - ICP-Fusion	ug/g	94.60	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	94.50	2.04e+03	n/a	n/a
S96T003891		F Samarium - ICP-Fusion	ug/g	98.80	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	99.10	1.02e+03	n/a	n/a
S96T003891		F Strontium - ICP-Fusion	ug/g	97.40	<1.00e-01	<2.04e+03	<2.06e3	n/a	n/a	93.70	2.04e+03	n/a	n/a
S96T003891		F Tantalum - ICP-Fusion	ug/g	96.40	<1.00e-02	<2.04e+02	<2.06e2	n/a	n/a	93.80	204.0	n/a	n/a
S96T003891		F Thorium - ICP-Fusion	ug/g	97.80	<1.00e-01	<2.04e+02	<2.06e2	n/a	n/a	95.90	204.0	n/a	n/a
S96T003891		F Uranium - ICP-Fusion	ug/g	99.40	<5.00e-02	<1.02e+04	<1.03e4	n/a	n/a	88.50	1.02e+04	n/a	n/a
S96T003891		F Vanadium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.02e+03	<1.03e3	n/a	n/a	96.80	1.02e+03	n/a	n/a
S96T003891		F Zinc - ICP-Fusion	ug/g	101.0	<1.00e-02	3.75e+02	375.0	375.0	0.00	97.90	204.0	n/a	n/a
S96T003891		F Zirconium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.04e+02	<2.06e2	n/a	n/a	94.60	204.0	n/a	n/a
S96T003891		F Alpha of Digested Solid	uCi/g	110.2	<1.80e-03	7.22e+03	7.75e+03	7.49e+03	7.08	84.12	4.00e+03	5.54e+01	n/a
S96T003904		A Silver - ICP-Acid Digest	ug/g	24.80	<1.00e-02	16.20	16.00	16.10	1.24	54.70	3.760	n/a	n/a
S96T003904		A Aluminum - ICP-Acid Digest	ug/g	101.2	<1.44e-01	2.07e+04	2.10e+04	2.08e+04	1.44	137.6	18.80	n/a	n/a
S96T003904		A Arsenic - ICP-Acid Digest	ug/g	98.00	<1.00e-01	<3.75e+03	<3.71e1	n/a	n/a	97.20	37.60	n/a	n/a
S96T003904		A Boron - ICP-Acid Digest	ug/g	121.2	<4.34e-01	66.10	66.05	66.05	0.15	101.2	18.80	n/a	n/a
S96T003904		A Barium - ICP-Acid Digest	ug/g	98.60	<5.00e-02	<1.88e+01	<1.85e1	n/a	n/a	96.40	18.80	n/a	n/a
S96T003904		A Beryllium - ICP-Acid Digest	ug/g	97.40	<5.00e-03	<1.88e0	<1.85e0	n/a	n/a	95.80	1.880	n/a	n/a

Sample#	R	Ad	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
5961003904	A	81 smch	-ICP-Acid Digest	ug/g	95.20	<1.00e-01	< 37.50	< 37.1e1	n/a	n/a	95.60	37.60	n/a	n/a
5961003904	A	Calcium	-ICP-Acid Digest	ug/g	100.2	2.320	4.08e+02	400.0	404.0	1.98	82.00	37.60	n/a	n/a
5961003904	A	Calcium	-ICP-Acid Digest	ug/g	93.80	<5.00e-03	7.000	7.010	7.005	0.14	95.80	1.880	n/a	n/a
5961003904	A	Cerium	-ICP-Acid Digest	ug/g	100.6	<1.00e-01	< 37.50	<3.71e1	n/a	n/a	98.80	37.60	n/a	n/a
5961003904	A	Cobalt	-ICP-Acid Digest	ug/g	96.80	<2.00e-02	7.500	<7.41e0	n/a	n/a	97.60	7.500	n/a	n/a
5961003904	A	Chromium	-ICP-Acid Digest	ug/g	95.20	<1.00e-02	1.56e+03	1.59e+03	1.58e+03	1.90	101.8	3.760	n/a	n/a
5961003904	A	Copper	-ICP-Acid Digest	ug/g	95.40	<1.00e-02	7.810	9.050	8.430	14.7	92.40	3.760	n/a	n/a
5961003904	A	Iron	-ICP-Acid Digest	ug/g	96.40	<7.00e-02	61.20	59.00	60.10	3.66	113.4	18.80	n/a	n/a
5961003904	A	Potassium	-ICP-Acid Digest	ug/g	98.40	<5.00e-02	3.26e+03	3.28e+03	3.28e+03	0.92	93.60	188.0	n/a	n/a
5961003904	A	Lanthanum	-ICP-Acid Digest	ug/g	100.4	<5.00e-02	< 18.80	<1.85e1	n/a	n/a	99.40	18.80	n/a	n/a
5961003904	A	Lithium	-ICP-Acid Digest	ug/g	99.40	<1.00e-02	< 3.750	<3.71e0	n/a	n/a	100.4	3.760	n/a	n/a
5961003904	A	Magnesium	-ICP-Acid Digest	ug/g	93.00	<1.00e-01	< 37.50	<3.71e1	n/a	n/a	88.80	37.60	n/a	n/a
5961003904	A	Manganese	-ICP-Acid Digest	ug/g	96.40	<5.00e-02	48.80	49.20	49.00	0.82	98.20	18.80	n/a	n/a
5961003904	A	Molybdenum	-ICP-Acid Digest	ug/g	138.6	<1.110	2.03e+05	1.99e+05	2.01e+05	1.99	214.0	37.60	n/a	n/a
5961003904	A	Sodium	-ICP-Acid Digest	ug/g	99.40	<1.00e-01	< 37.50	<3.71e1	n/a	n/a	97.40	37.60	n/a	n/a
5961003904	A	Neodymium	-ICP-Acid Digest	ug/g	96.60	<2.00e-02	28.20	25.80	27.00	8.89	96.40	7.500	n/a	n/a
5961003904	A	Nickel	-ICP-Acid Digest	ug/g	94.60	<2.00e-01	1.67e+03	581.0	1.13e+03	96.8	184.0	75.00	n/a	n/a
5961003904	A	Lead	-ICP-Acid Digest	ug/g	93.20	<1.00e-01	3.59e+03	3.43e+03	3.47e+03	1.17	106.6	37.60	n/a	n/a
5961003904	A	Phosphorus	-ICP-Acid Digest	ug/g	89.40	<5.00e-02	< 22.50	<2.22e1	n/a	n/a	93.20	22.60	n/a	n/a
5961003904	A	Sulfur	-ICP-Acid Digest	ug/g	89.20	<1.00e-01	< 37.50	<3.71e1	n/a	n/a	91.40	37.60	n/a	n/a
5961003904	A	Antimony	-ICP-Acid Digest	ug/g	264.0	<2.00e-01	1.17e+02	110.0	113.5	6.17	103.0	18.80	n/a	n/a
5961003904	A	Selenium	-ICP-Acid Digest	ug/g	97.20	<1.00e-01	< 37.50	<3.71e1	n/a	n/a	96.00	37.60	n/a	n/a
5961003904	A	Strontium	-ICP-Acid Digest	ug/g	97.20	<1.00e-02	< 3.750	<3.71e0	n/a	n/a	95.40	37.60	n/a	n/a
5961003904	A	Titanium	-ICP-Acid Digest	ug/g	95.20	<1.00e-02	< 3.750	<3.71e0	n/a	n/a	95.00	37.60	n/a	n/a
5961003904	A	Thallium	-ICP-Acid Digest	ug/g	89.80	<2.00e-01	< 75.00	<7.41e1	n/a	n/a	91.20	75.00	n/a	n/a
5961003904	A	Uranium	-ICP-Acid Digest	ug/g	95.40	<5.00e-01	<1.88e+02	<1.85e2	n/a	n/a	99.70	188.0	n/a	n/a
5961003904	A	Vanadium	-ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 18.80	<1.85e1	n/a	n/a	94.80	18.80	n/a	n/a
5961003904	A	Zinc	-ICP-Acid Digest	ug/g	90.20	<3.00e-02	38.20	35.60	36.90	7.05	93.20	37.60	n/a	n/a
5961003904	A	Zirconium	-ICP-Acid Digest	ug/g	96.40	<1.00e-02	7.450	7.720	7.585	3.56	97.00	3.760	n/a	n/a
5961003913	W	Acetate by IC-Dioxex	4000/4500	ug/g	105.6	<2.00e-02	8.61e+02	786.0	808.4	13.0	55.65	43.81	n/a	n/a
5961003913	W	Formate by IC-Dioxex	4000/4500	ug/g	95.83	<2.10e-02	1.02e+03	988.0	1.00e+03	3.19	91.67	46.00	n/a	n/a
5961003913	W	Fluoride-IC-Dioxex	4000/4500	ug/g	102.5	<1.30e-02	1.71e+03	1.89e+03	1.80e+03	10.0	73.56	54.37	n/a	n/a
5961003913	W	Chloride-IC-Dioxex	4000/4500	ug/g	105.2	<2.50e-02	5.28e+03	4.63e+03	4.96e+03	13.1	60.63	71.11	n/a	n/a
5961003913	W	Nitrate-IC - Dioxex	4000/4500	ug/g	102.9	<1.07e-01	6.63e+04	5.76e+04	6.19e+04	14.0	30.76	447.5	n/a	n/a
5961003913	W	Bromide by Ion Chromatograph		ug/g	103.4	<1.26e-01	<5.27e+02	<5.72e2	n/a	n/a	95.08	526.9	n/a	n/a
5961003913	W	Nitrate by IC-Dioxex	4000/4500	ug/g	104.0	<1.16e-01	1.95e+05	1.56e+05	1.56e+05	0.64	-6.450e1	585.1	n/a	n/a
5961003913	W	Phosphate-IC-Dioxex	4000/4500	ug/g	104.0	<1.19e-01	1.36e+03	3.86e+03	2.61e+03	95.8	97.25	497.3	n/a	n/a
5961003913	W	Sulfate by IC-Dioxex	4000/4500	ug/g	102.8	<1.36e-01	9.46e+03	9.06e+03	9.26e+03	4.32	91.60	568.5	n/a	n/a
5961003913	W	Oxalate by IC-Dioxex	4000/4500	ug/g	106.6	<1.05e-01	4.95e+03	4.49e+03	4.72e+03	9.75	93.57	439.1	n/a	n/a

Table 1. Data Summary Table
AN-105

CORE NUMBER: 152

SEGMENT #: Drainable Liquid Composite

SEGMENT PORTION: Core Composite

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004034		%	%									
S96T004039		Volume % Solids	Joules/g	94.55	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T004039		DSC Exotherm using Mettler	Joules/g	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T004039		DSC Exotherm Dry, Calculated	Joules/g	98.72	n/a	53.57	46.67	50.12	13.8	n/a	n/a	n/a
S96T004039		% Water by TGA using Mettler	%	98.72	n/a	53.57	46.67	50.12	13.8	n/a	n/a	n/a
S96T004039		Fluoride-IC-Dioxon 4000/4500	ug/mL	92.54	<1.30e-02	1.02e+04	<5.38e-1	1.02e+04	0.00	145.4	53.83	n/a
S96T004039		Chloride-IC-Dioxon 4000/4500	ug/mL	96.20	<1.70e-02	1.02e+04	1.02e+04	1.02e+04	0.00	92.41	70.40	n/a
S96T004039		Nitrite-IC - Dioxon 4000/4500	ug/mL	95.39	<1.07e-01	1.31e+05	1.30e+05	1.30e+05	0.77	89.85	443.1	n/a
S96T004039		Bromide by Ion Chromatograph	ug/mL	97.45	<2.26e-01	<5.22e-2	<5.22e-2	<5.22e-2	n/a	94.40	521.8	n/a
S96T004039		Nitrate by IC-Dioxon 4000/4500	ug/mL	96.30	<1.40e-01	1.68e+05	1.67e+05	1.67e+05	0.60	88.72	592.7	n/a
S96T004039		Phosphate-IC-Dioxon 4000/4500	ug/mL	96.70	<1.19e-01	6.35e+02	598.0	616.4	6.00	95.96	479.4	n/a
S96T004039		Sulfate by IC-Dioxon 4000/4500	ug/mL	96.20	<1.36e-01	1.01e+03	<4.35e+02	968.5	8.35	95.72	562.8	n/a
S96T004039		Oxalate by IC-Dioxon 4000/4500	ug/mL	101.7	<1.05e-01	<1.54e-02	<1.54e-2	n/a	n/a	101.7	434.8	n/a
S96T004039		Alphal in Liquid Samples	ug/mL	91.41	<5.00e-03	<7.00e-02	<0.070	n/a	n/a	79.67	7.80e-02	5.00e+02
S96T004049		Mercury by CVAAS (Perkin Elmer)	ug/mL	89.00	<5.00e-03	<7.00e-02	<0.070	n/a	n/a	95.80	205.0	n/a
S96T004049		Tot. Organic Carbon by Coul.	ug/mL	95.00	1.500	6.13e+03	6.07e+03	6.10e+03	0.98	95.80	205.0	n/a
S96T004049		TIC by Acid/Coulometry	ug/mL	92.35	10.80	2.22e+03	2.17e+03	2.20e+03	2.28	n/a	5.000	n/a
S96T004049		TOC by Persulfate/Coulometry	ug/mL	92.33	10.80	2.74e+03	2.44e+03	2.59e+03	11.6	121.0	40.00	n/a
S96T004049		Technetium-99 Lig. Scint.	ug/mL	102.7	<1.59e-05	2.79e-01	2.68e-01	2.74e-01	4.02	n/a	1.00e-03	8.90E-01
S96T004049		Strontium-89/90 High Level	ug/mL	94.26	<3.29e-05	4.20e-02	4.08e-02	4.14e-02	2.90	n/a	4.66e-05	1.02E+00
S96T004049		Pu-239/240 by TRU-SPEC Resin	ug/mL	94.49	<3.68e-05	<3.47e-05	<3.72E-5	n/a	n/a	n/a	3.47e-05	1.00E+02
S96T004049		Np237 by TTA Extraction	ug/mL	79.48	<2.06e-05	<2.06e-05	<2.21E-5	n/a	n/a	n/a	4.32e-05	5.00E+02
S96T004049		Iodine-129 Waste Tank Samples	ug/mL	81.93	<4.78e-05	2.46e-04	2.89e-04	2.67e-04	16.1	n/a	1.84e-05	4.9
S96T004049		Cobalt-60 by GEA	ug/mL	101.4	<4.65e-03	<1.19e-02	<1.30e-2	n/a	n/a	n/a	1.20e-02	n/a
S96T004049		Cesium-137 by GEA	ug/mL	102.0	<1.27e-02	4.73e+02	476.0	474.5	0.63	n/a	n/a	0.140
S96T004049		Europium-154 by GEA	ug/mL	n/a	<1.28e-02	<9.74e-02	<9.48e-2	n/a	n/a	n/a	9.70e-02	n/a
S96T004049		Europium-155 by GEA	ug/mL	n/a	<3.26e-01	<3.26e-1	<3.26e-1	n/a	n/a	n/a	3.25e-01	n/a
S96T004049		Americium-241 by GEA	ug/mL	n/a	<3.14e-02	<8.51e-01	<8.57e-1	n/a	n/a	n/a	8.51e-01	n/a
S96T004049		Beta in Liquid Samples	ug/mL	115.1	<8.77e-02	5.42e+02	534.0	538.0	1.49	99.34	1.11e-01	2.95E-01
S96T004049		Am-241 by Extraction	ug/mL	100.8	<3.96e-05	<3.91e-05	<3.52E-5	n/a	n/a	n/a	3.91e-05	8.65E+00
S96T004049		Am-243 by Extraction	ug/mL	100.8	<3.96e-05	<3.91e-05	<3.52E-5	n/a	n/a	n/a	3.91e-05	1.00E+02
S96T004049		Am-243/244 by Extraction	ug/mL	90.80	<1.00e-02	15.40	14.90	15.15	3.30	83.70	5.240	n/a
S96T004144		Silver -ICP-Acid Digest-Liquid	ug/mL	91.00	<1.24e-01	3.56e+04	3.59e+04	3.57e+04	0.56	108.8	26.20	n/a
S96T004144		Aluminum -ICP-Acid Digest-Liq	ug/mL	91.00	<1.00e-01	<5.240	<5.00e1	n/a	n/a	90.00	52.40	n/a
S96T004144		Arsenic -ICP-Acid Digest-Liq	ug/mL	93.20	<1.00e-01	<5.240	<5.00e1	n/a	n/a	90.60	26.20	n/a
S96T004144		Boron -ICP-Acid Digest-Liquid	ug/mL	102.2	<3.90e-01	1.56e+02	183.0	169.5	15.9	87.40	26.20	n/a
S96T004144		Barium -ICP-Acid Digest-Liquid	ug/mL	91.80	<5.00e-03	<2.620	<2.50e1	n/a	n/a	89.20	2.620	n/a
S96T004144		Beryllium -ICP-Acid Digest-Liq	ug/mL	94.60	<5.00e-03	<2.620	<2.50e1	n/a	n/a	89.20	2.620	n/a
S96T004144		Bismuth -ICP-Acid Digest-Liq	ug/mL	90.40	<1.00e-01	<5.240	<5.00e1	n/a	n/a	89.80	52.40	n/a
S96T004144		Calcium -ICP-Acid Digest-Liq	ug/mL	143.2	1.610	9.03e+02	853.0	878.0	5.69	67.80	52.40	n/a
S96T004144		Cadmium -ICP-Acid Digest-Liq	ug/mL	88.60	<5.00e-03	<2.620	<2.50e1	n/a	n/a	88.00	2.620	n/a
S96T004144		Cerium -ICP-Acid Digest-Liq	ug/mL	93.20	<1.00e-01	<5.240	<5.00e1	n/a	n/a	88.00	52.40	n/a
S96T004144		Cobalt -ICP-Acid Digest-Liq	ug/mL	91.20	<2.00e-02	<10.50	<9.99e0	n/a	n/a	89.40	10.50	n/a
S96T004144		Chromium -ICP-Acid Digest-Liq	ug/mL	89.60	<1.00e-02	1.87e+02	187.0	187.0	0.00	88.00	5.240	n/a
S96T004144		Copper -ICP-Acid Digest-Liquid	ug/mL	89.40	<1.00e-02	<5.240	6.290	n/a	n/a	85.60	5.240	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004144	B	Iron - ICP-Acid Digest-Liquid	ug/mL	90.40	<5.00e-02	72.50	55.10	63.80	27.3	84.40	26.20	n/a
S96T004144	B	Potassium - ICP-Acid Digest-Liquid	ug/mL	84.80	<5.00e-01	5.34e+03	5.27e+03	5.30e+03	1.32	85.40	262.0	n/a
S96T004144	B	Lanthanum - ICP-Acid Digest-Liquid	ug/mL	93.20	<1.00e-02	< 26.20	<2.50e1	n/a	n/a	89.60	26.20	n/a
S96T004144	B	Lithium - ICP-Acid Digest-Liquid	ug/mL	93.20	<1.00e-02	5.820	6.110	5.965	4.86	90.40	5.240	n/a
S96T004144	B	Magnesium - ICP-Acid Digest-Liquid	ug/mL	86.80	<1.00e-02	< 52.40	<5.00e1	n/a	n/a	84.20	52.40	n/a
S96T004144	B	Manganese - ICP-Acid Digest-Liquid	ug/mL	87.80	<1.00e-02	< 5.240	<5.00e0	n/a	n/a	85.00	5.240	n/a
S96T004144	B	Molybdenum - ICP-Acid Digest-Liquid	ug/mL	89.20	<5.00e-02	82.30	82.80	82.55	0.61	87.40	26.20	n/a
S96T004144	B	Sodium - ICP-Acid Digest-Liquid	ug/mL	111.6	1.110	2.11e+05	2.11e+05	2.11e+05	0.00	119.2	52.40	n/a
S96T004144	B	Neodymium - ICP-Acid Digest-Liquid	ug/mL	92.20	<1.00e-01	< 52.40	<5.00e1	n/a	n/a	88.20	52.40	n/a
S96T004144	B	Nickel - ICP-Acid Digest-Liquid	ug/mL	91.20	<2.00e-02	< 10.50	<9.99e0	n/a	n/a	89.60	10.50	n/a
S96T004144	B	Phosphorus - ICP-Acid Digest-Liquid	ug/mL	96.20	<2.00e-01	3.36e+02	343.0	339.5	2.06	95.20	105.0	n/a
S96T004144	B	Lead - ICP-Acid Digest-Liquid	ug/mL	86.20	<1.00e-01	< 52.40	<5.00e1	n/a	n/a	90.00	52.40	n/a
S96T004144	B	Sulfur - ICP-Acid Digest-Liquid	ug/mL	91.00	<1.00e-01	8.74e+02	880.0	877.0	0.68	93.20	52.40	n/a
S96T004144	B	Antimony - ICP-Acid Digest-Liquid	ug/mL	89.00	<6.00e-02	< 31.40	<3.00e1	n/a	n/a	88.40	31.40	n/a
S96T004144	B	Selenium - ICP-Acid Digest-Liquid	ug/mL	85.20	<1.00e-01	< 52.40	<5.00e1	n/a	n/a	86.40	52.40	n/a
S96T004144	B	Silicon - ICP-Acid Digest-Liquid	ug/mL	130.0	1.44e-01	2.08e+02	231.0	219.5	10.5	87.00	26.20	n/a
S96T004144	B	Samarium - ICP-Acid Digest-Liquid	ug/mL	90.40	<1.00e-01	< 52.40	<5.00e1	n/a	n/a	86.60	52.40	n/a
S96T004144	B	Titanium - ICP-Acid Digest-Liquid	ug/mL	90.60	<1.00e-02	< 5.240	<5.00e0	n/a	n/a	79.80	5.240	n/a
S96T004144	B	Strontium - ICP-Acid Digest-Liquid	ug/mL	83.00	<1.00e-02	< 5.240	<5.00e0	n/a	n/a	86.80	5.240	n/a
S96T004144	B	Thallium - ICP-Acid Digest-Liquid	ug/mL	85.00	<2.00e-01	<1.05e+02	<9.99e1	n/a	n/a	82.60	105.0	n/a
S96T004144	B	Uranium - ICP-Acid Digest-Liquid	ug/mL	88.10	<5.00e-01	<2.62e+02	<2.50e2	n/a	n/a	85.00	262.0	n/a
S96T004144	B	Vanadium - ICP-Acid Digest-Liquid	ug/mL	89.00	<1.00e-02	< 26.20	<2.50e1	n/a	n/a	86.80	26.20	n/a
S96T004144	B	Zinc - ICP-Acid Digest-Liquid	ug/mL	87.80	7.10e-02	77.70	74.50	76.10	4.20	86.20	5.240	n/a
S96T004144	B	Zirconium - ICP-Acid Digest-Liquid	ug/mL	90.60	<1.00e-02	< 5.240	<5.00e0	n/a	n/a	87.40	5.240	n/a
S96T004831	B	Chromium (VI) by Spec.	ug/mL	105.6	<2.20e-01	1.54e+02	153.0	153.5	0.65	108.8	22.18	n/a
S96T004831	D	OH- by Pot. Titration	ug/mL	101.0	23.90	5.66e+04	6.05e+04	5.86e+04	6.66	n/a	2.50e+03	n/a
S96T004831	D	Silver - ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	18.50	17.50	18.00	5.56	92.50	6.010	n/a
S96T004831	D	Aluminum - ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	4.34e+04	4.38e+04	4.36e+04	0.92	130.0	30.10	n/a
S96T004831	D	Arsenic - ICP-Acid Dil.	ug/mL	104.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	97.00	60.10	n/a
S96T004831	D	Boron - ICP-Acid Dil.	ug/mL	99.00	<5.00e-02	74.20	71.30	72.75	3.99	96.40	30.10	n/a
S96T004831	D	Barium - ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	97.70	30.10	n/a
S96T004831	D	Beryllium - ICP-Acid Dil.	ug/mL	103.6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	101.0	3.000	n/a
S96T004831	D	Bismuth - ICP-Acid Dil.	ug/mL	100.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	101.0	60.10	n/a
S96T004831	D	Calcium - ICP-Acid Dil.	ug/mL	104.1	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	106.0	60.10	n/a
S96T004831	D	Cadmium - ICP-Acid Dil.	ug/mL	101.0	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	100.0	3.000	n/a
S96T004831	D	Cerium - ICP-Acid Dil.	ug/mL	99.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	99.00	60.10	n/a
S96T004831	D	Cobalt - ICP-Acid Dil.	ug/mL	101.8	<1.00e-02	< 12.00	<1.20e1	n/a	n/a	100.0	12.00	n/a
S96T004831	D	Chromium - ICP-Acid Dil.	ug/mL	102.4	<1.00e-02	2.21e+02	224.0	222.5	1.35	101.0	6.010	n/a
S96T004831	D	Copper - ICP-Acid Dil.	ug/mL	103.1	<5.00e-02	< 6.010	<6.01e0	n/a	n/a	100.0	6.010	n/a
S96T004831	D	Iron - ICP-Acid Dil.	ug/mL	102.4	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a
S96T004831	D	Potassium - ICP-Acid Dil.	ug/mL	96.40	<5.00e-01	6.49e+03	6.55e+03	6.52e+03	0.92	144.0	300.0	n/a
S96T004831	D	Lanthanum - ICP-Acid Dil.	ug/mL	101.2	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a
S96T004831	D	Lithium - ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	101.0	6.010	n/a
S96T004831	D	Magnesium - ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	95.00	60.10	n/a
S96T004831	D	Manganese - ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	95.60	6.010	n/a
S96T004831	D	Molybdenum - ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	1.00e+02	102.0	101.0	1.98	101.0	30.10	n/a
S96T004831	D	Sodium - ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	2.54e+05	2.57e+05	2.56e+05	1.17	176.0	60.10	n/a
S96T004831	D	Strontium - ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	100.0	60.10	n/a
S96T004831	D	Neodymium - ICP-Acid Dil.	ug/mL	101.0	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	98.80	12.00	n/a
S96T004831	D	Nickel - ICP-Acid Dil.	ug/mL	99.20	<2.00e-01	3.90e+02	395.0	392.5	1.27	98.40	120.0	n/a
S96T004831	D	Phosphorus - ICP-Acid Dil.	ug/mL									

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004831	D	Lead-ICP-Acid Dil.	ug/mL	101.2	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	110.0	60.10	n/a
S96T004831	D	Sulfur-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	1.03e+03	993.0	1.01e+03	3.66	97.20	60.10	n/a
S96T004831	D	Antimony-ICP-Acid Dil.	ug/mL	96.40	<6.00e-02	< 36.10	<3.61e1	n/a	n/a	95.60	36.10	n/a
S96T004831	D	Selenium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	106.0	60.10	n/a
S96T004831	D	Silicon-ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	1.66e+02	166.0	166.0	0.00	99.40	30.10	n/a
S96T004831	D	Samarium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	96.60	60.10	n/a
S96T004831	D	Strontium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	97.50	6.010	n/a
S96T004831	D	Titanium-ICP-Acid Dil.	ug/mL	98.00	<2.00e-01	< 6.010	<6.01e0	n/a	n/a	96.40	120.0	n/a
S96T004831	D	Uranium-ICP-Acid Dil.	ug/mL	96.60	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	95.00	300.0	n/a
S96T004831	D	Vanadium-ICP-Acid Dil.	ug/mL	95.90	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	99.40	30.10	n/a
S96T004831	D	Zinc-ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	102.0	6.010	n/a
S96T004831	D	Vanadium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	17.10	18.10	17.60	5.68	98.50	6.010	n/a
S96T004831	D	Zirconium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	98.50	6.010	n/a
S96T004831	D	Acetate by IC-Dionex 4000/4500	ug/mL	96.77	<2.00e-02	6.41e+02	642.0	641.5	0.16	n/a	82.82	n/a
S96T004831	D	Formate by Ion Chromatograph	ug/mL	98.33	<2.10e-02	1.62e+03	1.62e+03	1.62e+03	0.00	n/a	86.96	n/a
S96T004831	D	Tritium By Lachat	uCi/mL	98.01	8.13e-05	<3.96e-05	<3.90E-5	n/a	n/a	88.98	3.96e-05	5.75E+00
S96T004831	D	% Ct. Error	% Ct. Error	1.09	5.300	5.750	5.830	5.790	5.83	n/a	1.00e-04	n/a
S96T005546	D	Specific Gravity	Sp.G.	98.71	n/a	1.422	1.408	1.415	0.99	n/a	1.00e-02	n/a
S96T005546	D	Uranium by Phosphorescence	ug/mL	94.21	3.35e-01	3.760	3.370	3.565	10.9	91.62	4.00e-03	1.75E+00
S96T005546	D	U-Phosphorescence Inst. Error	% Inst Error	3.94	1.410	1.750	2.450	2.100	2.45	n/a	n/a	n/a

CORE NUMBER: 152
SEGMENT #: 14-20

SEGMENT PORTION: Core Composite

Table 1. Data Summary Table
AN-105

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961004254		DSC Exotherm using Mettler	Joules/g	99.82	n/a	43.70	31.70	31.70	31.8	n/a	n/a	n/a	n/a
S961004254		DSC Exotherm DRY Calculated	Joules/g Dry	n/a	n/a	80.50	58.40	69.45	31.8	n/a	n/a	n/a	n/a
S961004254		% Water by IGA using Mettler	%	99.80	n/a	46.19	45.29	45.74	1.97	n/a	n/a	n/a	n/a
S961004254		TIC by Acid/Coulometry	ug/g	100.2	2.500	1.25e+04	1.30e+04	1.30e+04	8.43	76.50	5.000	n/a	n/a
S961004254		TIC by Persulfate/Coulometry	ug/g	98.33	7.00e-01	4.90e+03	5.42e+03	5.42e+03	19.4	92.20	40.00	n/a	n/a
S961004254		F Uranium by Phosphorescence	ug/g	106.7	<1.820	88.60	86.20	87.40	2.75	74.34	1.820	1.31E+00	n/a
S961004254		F Technetium-99 Lig. Scint.	ug/g	105.3	<1.51e-02	1.81e-01	1.80e-01	1.80e-01	0.55	n/a	1.50e-02	4.42E+00	n/a
S961004254		F Strontium-87/90 High Level	ug/g	94.29	8.90e-02	9.410	8.910	9.160	5.66	n/a	1.20e-02	1.42E+00	n/a
S961004254		F Pu-239/240 by IR-SPEC Resin	ug/g	96.64	<1.69e-02	1.81e-03	<1.78e-3	n/a	n/a	n/a	2.00e-03	7.75E+00	n/a
S961004254		F Pu-237 by TIA Extraction	ug/g	72.99	<1.51e-02	3.75e-02	<2.98e-2	1.06e-02	5.68	n/a	1.90e-02	1.11E+02	n/a
S961004254		F Pu-237 by TIA Extraction	ug/g	99.83	<1.51e-02	3.75e-02	<2.98e-2	304.1	0.68	n/a	3.80e-02	n/a	n/a
S961004254		F Cesium-137 by GE	ug/g	101.8	<3.99e-02	<8.70e-02	<9.02e-2	303.0	n/a	n/a	n/a	0.290	n/a
S961004254		F Europium-154 by GE	ug/g	n/a	<3.46e-02	<8.25e-01	<8.22e-1	n/a	n/a	n/a	8.70e-02	n/a	n/a
S961004254		F Americium-241 by GE	ug/g	n/a	<1.78e-02	<8.25e-01	<8.22e-1	n/a	n/a	n/a	3.87e-01	n/a	n/a
S961004254		F Americium-241 by Extraction	ug/g	84.87	<1.49e-04	4.95e-03	3.24e-03	4.09e-03	41.8	n/a	1.00e-03	2.26E+00	n/a
S961004254		F Uran-233/234 by Extraction	ug/g	121.9	9.00e-03	1.52e-02	1.20e-02	1.36e-02	23.5	84.40	1.60e-02	9.62E+01	n/a
S961004254		F Alpha or Digested Solid	ug/g	111.7	1.15e-01	3.28e+02	325.0	326.5	0.92	86.13	4.90e-02	2.70E+01	n/a
S961004254		F Silver - ICP-Acid Digest	ug/g	86.10	<1.00e-02	12.30	11.50	11.90	6.72	85.40	3.820	n/a	n/a
S961004257		A Aluminum - ICP-Acid Digest	ug/g	86.60	8.70e-02	2.02e+04	2.07e+04	2.04e+04	2.42	80.10	19.10	n/a	n/a
S961004257		A Arsenic - ICP-Acid Digest	ug/g	88.20	<1.00e-01	<36.30	<4.03e1	116.5	19.7	88.60	38.20	n/a	n/a
S961004257		A Barium - ICP-Acid Digest	ug/g	96.60	3.56e-01	1.05e+02	128.0	n/a	n/a	87.40	19.10	n/a	n/a
S961004257		A Beryllium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	<19.10	<2.02e1	n/a	n/a	84.40	19.10	n/a	n/a
S961004257		A Bromine - ICP-Acid Digest	ug/g	84.60	<1.00e-01	<38.30	<4.03e1	n/a	n/a	87.60	38.20	n/a	n/a
S961004257		A Calcium - ICP-Acid Digest	ug/g	95.00	3.12e-01	2.09e+02	222.0	215.5	6.03	87.00	38.20	n/a	n/a
S961004257		A Cadmium - ICP-Acid Digest	ug/g	84.80	<5.00e-03	8.130	6.660	7.295	22.9	85.00	1.910	n/a	n/a
S961004257		A Cerium - ICP-Acid Digest	ug/g	88.20	<1.00e-01	<38.30	<4.03e1	n/a	n/a	86.20	38.20	n/a	n/a
S961004257		A Cobalt - ICP-Acid Digest	ug/g	87.00	<2.00e-02	<7.660	<8.06e0	n/a	n/a	86.60	7.660	n/a	n/a
S961004257		A Chromium - ICP-Acid Digest	ug/g	85.40	<1.00e-02	1.60e+03	1.37e+03	1.48e+03	15.9	82.80	3.820	n/a	n/a
S961004257		A Copper - ICP-Acid Digest	ug/g	85.00	<5.00e-02	4.880	5.340	5.110	9.00	72.00	3.820	n/a	n/a
S961004257		A Iron - ICP-Acid Digest	ug/g	85.60	<5.00e-02	71.70	50.40	61.15	34.5	82.00	19.10	n/a	n/a
S961004257		A Potassium - ICP-Acid Digest	ug/g	81.40	<5.00e-01	3.07e+03	3.11e+03	3.09e+03	1.29	86.80	19.10	n/a	n/a
S961004257		A Lanthanum - ICP-Acid Digest	ug/g	88.20	<5.00e-02	<19.10	<2.02e1	n/a	n/a	85.60	38.20	n/a	n/a
S961004257		A Lithium - ICP-Acid Digest	ug/g	81.80	<1.00e-01	<38.30	<4.03e1	8.700	2.50	87.60	38.20	n/a	n/a
S961004257		A Magnesium - ICP-Acid Digest	ug/g	81.80	<1.00e-02	8.600	8.800	8.700	2.50	87.60	38.20	n/a	n/a
S961004257		A Manganese - ICP-Acid Digest	ug/g	84.40	<5.00e-02	8.460	7.180	7.830	16.4	81.60	38.20	n/a	n/a
S961004257		A Molybdenum - ICP-Acid Digest	ug/g	84.20	<5.00e-02	45.80	45.90	45.85	0.22	84.40	19.10	n/a	n/a
S961004257		A Sodium - ICP-Acid Digest	ug/g	103.2	6.57e-01	1.57e+05	1.53e+05	1.55e+05	2.58	56.20	38.20	n/a	n/a
S961004257		A Neodymium - ICP-Acid Digest	ug/g	86.80	<1.00e-01	<38.30	<4.03e1	n/a	n/a	84.80	38.20	n/a	n/a
S961004257		A Nickel - ICP-Acid Digest	ug/g	85.80	<2.00e-02	34.90	28.60	31.75	19.8	86.20	7.660	n/a	n/a
S961004257		A Phosphorus - ICP-Acid Digest	ug/g	85.80	<2.00e-01	4.18e+02	586.0	402.0	7.56	90.00	76.60	n/a	n/a

WHC-SD-WM-DP-199, REV. 1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004257	A	Lead - ICP-Acid Digest	ug/g	82.40	<1.00e-01	< 36.30	<4.03e1	n/a	n/a	86.80	38.20	n/a
S96T004257	A	Sulfur - ICP-Acid Digest	ug/g	84.80	<1.00e-01	< 32.00	2.03e+03	2.18e+03	13.3	69.80	38.20	n/a
S96T004257	A	Antimony - ICP-Acid Digest	ug/g	84.60	<6.00e-02	< 23.00	<2.42e1	n/a	n/a	85.80	23.00	n/a
S96T004257	A	Selenium - ICP-Acid Digest	ug/g	81.60	<1.00e-01	< 38.30	<4.03e1	n/a	n/a	82.80	38.20	n/a
S96T004257	A	Silicon - ICP-Acid Digest	ug/g	144.0	1.13e-01	1.39e+02	218.0	178.5	44.3	126.6	19.10	n/a
S96T004257	A	Samarium - ICP-Acid Digest	ug/g	85.40	<1.00e-01	< 38.30	<4.03e1	n/a	n/a	83.80	38.20	n/a
S96T004257	A	Strontium - ICP-Acid Digest	ug/g	85.80	<1.00e-02	< 3.830	<4.03e0	n/a	n/a	84.00	38.20	n/a
S96T004257	A	Titanium - ICP-Acid Digest	ug/g	78.40	<1.00e-02	< 3.830	<4.03e0	n/a	n/a	77.20	38.20	n/a
S96T004257	A	Thallium - ICP-Acid Digest	ug/g	81.40	<2.00e-01	< 76.60	<8.06e0	n/a	n/a	79.40	76.60	n/a
S96T004257	A	Uranium - ICP-Acid Digest	ug/g	83.30	<5.00e-01	<1.91e+02	<2.02e2	n/a	n/a	84.80	191.0	n/a
S96T004257	A	Vanadium - ICP-Acid Digest	ug/g	84.60	<5.00e-02	< 19.10	<2.02e1	n/a	n/a	83.60	38.20	n/a
S96T004257	A	Zinc - ICP-Acid Digest	ug/g	82.40	<1.00e-02	7.780	8.580	8.180	9.78	83.60	38.20	n/a
S96T004257	A	Zirconium - ICP-Acid Digest	ug/g	86.80	<1.00e-02	4.510	<4.03e0	n/a	n/a	85.80	38.20	n/a
S96T004258	M	Chromium (VI) by Spec.	ug/g	105.6	<2.20e-01	1.18e+02	120.0	119.0	1.68	106.8	46.69	n/a
S96T004258	M	OH- by Pot. Titration	ug/g	97.36	<63.00	2.36e+04	2.45e+04	2.40e+04	3.74	n/a	1.33e+04	n/a
S96T004258	M	Acetate by IC-DIONEX 4000/4500	ug/g	80.08	<2.00e-02	2.36e+02	474.0	472.1	0.85	107.3	89.31	n/a
S96T004258	M	Formate by IC-DIONEX 4000/4500	ug/g	92.30	<2.10e-02	1.22e+03	1.23e+03	1.22e+03	0.82	100.0	183.1	n/a
S96T004258	M	Fluoride by IC-DIONEX 4000/4500	ug/g	103.4	<1.20e-02	1.35e+03	929.0	1.14e+03	36.9	79.32	53.59	n/a
S96T004258	M	Chloride by IC-DIONEX 4000/4500	ug/g	102.2	<1.70e-02	5.85e+03	5.91e+03	5.88e+03	1.02	109.0	75.92	n/a
S96T004258	M	Nitrite by Ion Chromatograph	ug/g	100.9	<1.08e-01	7.29e+04	7.32e+04	7.30e+04	0.41	104.2	482.2	n/a
S96T004258	M	Bromide by Ion Chromatograph	ug/g	99.15	<1.25e-01	<5.58e+02	<5.38e2	n/a	n/a	96.10	558.2	n/a
S96T004258	M	Nitrate by IC-DIONEX 4000/4500	ug/g	103.5	2.21e-01	1.17e+05	1.09e+05	1.13e+05	7.08	97.31	620.8	n/a
S96T004258	M	Phosphate by IC-DIONEX 4000/4500	ug/g	99.45	<1.20e-01	1.28e+03	1.13e+03	1.21e+03	12.4	101.3	535.9	n/a
S96T004258	M	Sulfate by IC-DIONEX 4000/4500	ug/g	102.9	<1.38e-01	7.22e+03	7.73e+03	7.47e+03	6.82	103.6	616.1	n/a
S96T004258	M	Oxalate by IC-DIONEX 4000/4500	ug/g	101.7	<1.05e-01	6.29e+03	6.69e+03	6.49e+03	6.16	101.7	468.9	n/a
S96T004258	M	Iodine-129 Waste Tank Samples	uCi/g	113.1	<5.40e-03	<3.37e-03	<3.92E-3	n/a	n/a	94.82	1.00e-03	0.00E+00
S96T004258	M	Tritium by Lachat	uCi/g	93.89	2.00e-03	2.41e-03	1.99e-03	2.20e-03	19.1	94.82	1.00e-03	5.01E+00
S96T004258	M	R3 - Rel. % Count Error	% Ct. Error	1.08	5.230	5.010	5.140	5.075	5.14	n/a	1.00e-04	n/a

CORE NUMBER: 152
SEGMENT #: S14-20 Core Composite

SEGMENT PORTION: Core Composite

Table 1. Data Summary Table
AN-105

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S96T004207			Bulk Density of Sample	g/mL	n/a	n/a	1.560	n/a	n/a	n/a	n/a	5.00e-01		n/a

CORE NUMBER: 152
SEGMENT #: 22

SEGMENT PORTION: Core Composite

Table 1. Data Summary Table
AN-105

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004259		DSC Exotherm using Mettler	Joules/g	99.82	n/a	60.40	54.10	57.25	11.0	n/a	n/a	n/a
S96T004259		PSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	1.06e-02	94.50	100.2	11.5	n/a	n/a	n/a
S96T004259		% Water by TGA using Mettler	%	99.80	n/a	42.61	42.87	42.74	0.61	n/a	n/a	n/a
S96T004259		TIC by Acid/Coulometry	ug/g	100.2	2.500	8.51e+03	7.99e+03	8.25e+03	6.30	92.60	5.000	n/a
S96T004259		TIC by Persulfate/Coulometry	ug/g	98.33	7.00e-01	4.33e+03	4.63e+03	4.48e+03	6.70	79.30	40.00	n/a
S96T004259		TOC by Phosphorescence	ug/g	106.7	<1.820	84.30	80.60	82.45	4.49	76.10	1.720	1.33e+00
S96T004261		F Technium-99 Lig. Scint.	uCi/g	105.3	8.10e-02	1.42e-01	1.26e-01	1.34e-01	11.9	n/a	1.20e-02	4.68e+00
S96T004261		F Strontium-89/90 High Level	uCi/g	94.29	5.90e-02	35.80	37.30	36.55	4.10	n/a	1.40e-02	7.37e-01
S96T004261		F Pu-239/240 by TRU-SPEC Resin	uCi/g	96.64	<1.69e-03	7.52e-03	7.57e-03	7.54e-03	0.66	n/a	2.00e-03	2.95e+00
S96T004261		F Pu-237 by TIA Extraction	uCi/g	72.99	<1.80e-02	<1.53e-02	<1.62e-02	n/a	n/a	n/a	1.80e-03	1.66e+02
S96T004261		F Cobalt-60 by GEA	uCi/g	99.83	<1.51e-02	<1.73e-02	<1.80e-02	n/a	n/a	n/a	1.70e-02	n/a
S96T004261		F Cesium-137 by GEA	uCi/g	101.8	4.50e-02	2.05e+02	207.0	206.0	0.97	n/a	n/a	0.240
S96T004261		F Europium-154 by GEA	uCi/g	n/a	<3.99e-02	<6.58e-02	<7.50e-02	n/a	n/a	n/a	6.60e-02	n/a
S96T004261		F Americium-155 by GEA	uCi/g	n/a	<3.46e-02	<2.21e-01	<2.30e-01	n/a	n/a	n/a	2.21e-01	n/a
S96T004261		F Americium-241 by GEA	uCi/g	n/a	<7.78e-02	<4.76e-01	<4.95e-01	n/a	n/a	n/a	4.76e-01	n/a
S96T004261		F Am-241 by Extraction	uCi/g	84.87	<7.49e-04	4.33e-02	3.45e-02	3.89e-02	22.6	n/a	6.00e-03	1.64e+00
S96T004261		F Cm-243/244 by Extraction	uCi/g	n/a	<7.49e-04	<5.63e-03	<4.50e-3	n/a	n/a	n/a	6.00e-03	4.95e+00
S96T004261		F Alpha of Digested Solid	uCi/g	121.9	9.00e-03	6.59e-02	5.00e-02	5.80e-02	27.4	100.3	1.50e-02	2.85e+01
S96T004261		F Beta of Solid Sample	uCi/g	111.7	1.15e-01	2.68e+02	281.0	274.5	4.74	101.3	4.70e-02	2.91e-01
S96T004262		A Silver -ICP-Acid Digest	ug/g	86.10	<1.00e-02	44.80	40.10	42.45	11.1	120.0	3.900	n/a
S96T004262		A Aluminum -ICP-Acid Digest	ug/g	86.60	8.70e-02	1.52e+04	1.37e+04	1.44e+04	10.4	150.8	19.50	n/a
S96T004262		A Arsenic -ICP-Acid Digest	ug/g	88.20	<1.00e-01	<36.90	<3.70e1	n/a	n/a	125.6	39.00	n/a
S96T004262		A Boron -ICP-Acid Digest	ug/g	96.60	3.50e-01	1.36e+02	91.20	113.6	39.4	120.6	19.50	n/a
S96T004262		A Barium -ICP-Acid Digest	ug/g	86.80	<5.00e-02	<19.50	<1.85e1	n/a	n/a	121.6	19.50	n/a
S96T004262		A Bismuth -ICP-Acid Digest	ug/g	91.80	<5.00e-03	<1.950	<1.85e0	n/a	n/a	126.0	1.950	n/a
S96T004262		A Calcium -ICP-Acid Digest	ug/g	95.00	3.12e-01	3.55e+02	302.0	328.5	16.1	119.6	39.00	n/a
S96T004262		A Cadmium -ICP-Acid Digest	ug/g	84.60	<1.00e-01	<36.90	<3.70e1	n/a	n/a	121.4	39.00	n/a
S96T004262		A Cerium -ICP-Acid Digest	ug/g	84.80	<5.00e-03	14.40	13.30	13.85	7.94	120.2	1.950	n/a
S96T004262		A Cobalt -ICP-Acid Digest	ug/g	88.20	<1.00e-01	<36.90	<3.70e1	n/a	n/a	122.2	7.780	n/a
S96T004262		A Chromium -ICP-Acid Digest	ug/g	85.40	<1.00e-02	2.90e+03	2.58e+03	2.74e+03	11.7	118.0	3.900	n/a
S96T004262		A Copper -ICP-Acid Digest	ug/g	85.00	<1.00e-02	7.040	5.020	6.030	33.5	118.2	3.900	n/a
S96T004262		A Iron -ICP-Acid Digest	ug/g	81.40	<5.00e-01	1.54e+02	2.05e+03	1.46e+3	10.2	119.2	19.50	n/a
S96T004262		A Lanthanum -ICP-Acid Digest	ug/g	81.40	<5.00e-01	2.24e+03	2.05e+03	2.14e+3	8.86	125.2	19.50	n/a
S96T004262		A Lithium -ICP-Acid Digest	ug/g	88.00	<1.00e-02	<19.50	<1.85e1	n/a	n/a	124.0	19.50	n/a
S96T004262		A Magnesium -ICP-Acid Digest	ug/g	81.80	<1.00e-01	7.320	22.10	14.71	100	126.2	3.900	n/a
S96T004262		A Manganese -ICP-Acid Digest	ug/g	83.40	<1.00e-02	<36.90	<3.70e1	n/a	n/a	118.2	39.00	n/a
S96T004262		A Molybdenum -ICP-Acid Digest	ug/g	84.20	<5.00e-02	52.20	46.60	49.40	11.3	116.4	3.900	n/a
S96T004262		A Nickel -ICP-Acid Digest	ug/g	103.2	6.57e-01	37.40	33.40	35.40	12.8	118.6	19.50	n/a
S96T004262		A Neodymium -ICP-Acid Digest	ug/g	86.80	<1.00e-01	1.75e+05	1.54e+05	1.64e+5	12.8	16.68	39.00	n/a
S96T004262		A Niobium -ICP-Acid Digest	ug/g	87.60	<2.00e-02	<36.90	<3.70e1	n/a	n/a	121.6	39.00	n/a
S96T004262		A Phosphorus -ICP-Acid Digest	ug/g	85.80	<2.00e-01	1.50e+04	1.32e+04	1.41e+4	12.8	122.2	7.780	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961004262	A	Lead -ICP-Acid Digest	ug/g	82.40	<1.00e-01	51.80	47.60	49.70	8.45	117.0	39.00	n/a
S961004262	A	Sulfur -ICP-Acid Digest	ug/g	84.80	<1.00e-01	1.53e+03	1.36e+03	1.44e+03	11.8	120.4	39.00	n/a
S961004262	A	Antimony -ICP-Acid Digest	ug/g	84.60	<5.00e-02	<23.40	<2.22e1	n/a	n/a	121.8	23.40	n/a
S961004262	A	Selenium -ICP-Acid Digest	ug/g	81.60	<1.00e-01	<38.90	<3.70e1	n/a	n/a	116.2	39.00	n/a
S961004262	A	Silicon -ICP-Acid Digest	ug/g	85.40	1.13e-01	6.21e+02	328.0	474.5	61.7	136.2	19.50	n/a
S961004262	A	Samarium -ICP-Acid Digest	ug/g	85.40	<1.00e-01	<3.890	<3.70e1	n/a	n/a	119.8	39.00	n/a
S961004262	A	Strontium -ICP-Acid Digest	ug/g	85.80	<1.00e-02	<3.890	<3.70e1	n/a	n/a	120.0	3.900	n/a
S961004262	A	Titanium -ICP-Acid Digest	ug/g	78.40	<1.00e-02	<3.890	<3.70e1	n/a	n/a	109.8	3.900	n/a
S961004262	A	Thallium -ICP-Acid Digest	ug/g	81.40	<2.00e-01	<77.90	<7.40e1	n/a	n/a	113.6	77.80	n/a
S961004262	A	Vanadium -ICP-Acid Digest	ug/g	83.30	<5.00e-01	<1.95e+02	<1.85e2	n/a	n/a	119.4	19.50	n/a
S961004262	A	Zinc -ICP-Acid Digest	ug/g	84.60	<5.00e-02	<19.50	<1.85e1	n/a	n/a	117.4	3.900	n/a
S961004262	A	Zirconium -ICP-Acid Digest	ug/g	82.40	<1.00e-02	53.70	59.00	55.85	11.3	40.20	3.900	n/a
S961004262	A	Chromium (VI) by Spec.	ug/g	86.80	<1.00e-02	1.07e+02	101.0	104.0	5.77	106.5	41.31	n/a
S961004263	W	OH - by Pot. Titration	ug/g	97.36	<63.00	2.24e+04	2.16e+04	2.20e+04	3.64	n/a	1.18e+04	n/a
S961004263	W	Acetate by IC-DIONEX 4000/4500	ug/g	80.08	<2.00e-02	<79.01	<7.83e1	n/a	n/a	n/a	79.00	n/a
S961004263	W	Formate by IC-DIONEX 4000/4500	ug/g	92.50	<2.10e-02	1.12e+03	1.11e+03	1.12e+03	0.90	96.67	162.0	n/a
S961004263	W	Fluoride-IC-DIONEX 4000/4500	ug/g	103.4	<1.20e-02	5.18e+03	4.87e+03	5.02e+03	6.17	96.95	67.16	n/a
S961004263	W	Chloride-IC-DIONEX 4000/4500	ug/g	102.2	<1.20e-02	4.34e+03	3.80e+03	4.07e+03	13.3	101.3	67.16	n/a
S961004263	W	Nitrite-IC - Dionex 4000/4500	ug/g	100.9	<1.28e-01	5.11e+04	4.43e+04	4.77e+04	14.3	103.5	426.7	n/a
S961004263	W	Bromide by Ion Chromatograph	ug/g	99.15	<1.28e-01	<4.94e+02	<4.89e2	n/a	n/a	99.76	493.7	n/a
S961004263	W	Nitrate by IC-DIONEX 4000/4500	ug/g	103.5	2.21e-01	1.75e+05	1.47e+05	1.61e+05	17.4	84.18	549.7	n/a
S961004263	W	Phosphate-IC-DIONEX 4000/4500	ug/g	99.45	<1.28e-01	4.80e+04	4.34e+04	4.57e+04	10.1	102.0	474.0	n/a
S961004263	W	Sulfate by IC-DIONEX 4000/4500	ug/g	102.9	<1.38e-01	4.40e+03	3.96e+03	4.18e+03	10.5	102.2	545.2	n/a
S961004263	W	Oxalate by IC-DIONEX 4000/4500	ug/g	101.7	<1.05e-01	6.30e+03	5.46e+03	5.88e+03	14.3	101.9	414.8	n/a
S961004263	W	Iodine-129 Waste Tank Samples	uCi/g	113.1	<5.40e-03	<3.03e-03	<3.11E-3	n/a	n/a	n/a	3.00e-03	0.00E+00
S961004263	W	Trilium By Lachat	uCi/g	94.66	<6.38e-04	<6.33e-04	<6.28E-4	n/a	n/a	n/a	1.00e-03	5.89E+00
S961004263	W	H3 - Ret. % Count Error	% Ct. Error	1.11	6.430	5.890	5.830	5.860	5.83	n/a	1.00e-04	n/a

U Upper Half of Segment: U Upper Half of Segment

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003712	A	Bulk Density of Sample	g/mL	n/a	n/a	1.640	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961003798	DSC	Exotherm Using Mettler	Joules/g	98.77	n/a	59.00	73.00	66.00	21.2	n/a	n/a	n/a
S961003798	DSC	Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	98.10	121.0	109.5	20.9	n/a	n/a	n/a
S961003798	%	Water by TGA Using Mettler	%	99.51	n/a	38.38	40.86	39.87	4.97	n/a	n/a	n/a
S961003892	F	Silver -ICP-Fusion	ug/g	99.00	<1.00e-02	<2.09e+02	<2.12e2	2.18e+04	n/a	86.50	209.0	n/a
S961003892	F	Aluminum -ICP-Fusion	ug/g	96.20	<1.00e-02	2.18e+04	2.19e+04	2.18e+04	0.46	93.40	1.05e+03	n/a
S961003892	F	Arsenic -ICP-Fusion	ug/g	102.0	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	101.0	2.09e+03	n/a
S961003892	F	Boron -ICP-Fusion	ug/g	98.00	<5.00e-02	<1.04e+03	<1.06e3	n/a	n/a	94.40	1.05e+03	n/a
S961003892	F	Barium -ICP-Fusion	ug/g	98.20	<5.00e-02	<1.04e+03	<1.06e3	n/a	n/a	94.20	1.05e+03	n/a
S961003892	F	Beryllium -ICP-Fusion	ug/g	100.4	<5.00e-03	<1.04e+02	<1.06e2	n/a	n/a	96.70	105.0	n/a
S961003892	F	Bismuth -ICP-Fusion	ug/g	101.4	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	99.20	2.09e+03	n/a
S961003892	F	Calcium -ICP-Fusion	ug/g	100.8	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	101.0	2.09e+03	n/a
S961003892	F	Cadmium -ICP-Fusion	ug/g	100.2	<1.00e-03	<1.04e+02	<1.06e2	n/a	n/a	98.00	105.0	n/a
S961003892	F	Cerium -ICP-Fusion	ug/g	98.60	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	96.30	2.09e+03	n/a
S961003892	F	Cobalt -ICP-Fusion	ug/g	100.6	<1.00e-02	<4.18e+02	<4.25e2	n/a	n/a	97.60	418.0	n/a
S961003892	F	Chromium -ICP-Fusion	ug/g	100.8	<1.00e-02	1.58e+03	1.62e+03	1.60e+03	2.50	98.40	209.0	n/a
S961003892	F	Copper -ICP-Fusion	ug/g	101.0	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	97.80	209.0	n/a
S961003892	F	Iron -ICP-Fusion	ug/g	100.0	<5.00e-02	<1.04e+03	<1.06e3	n/a	n/a	97.90	1.05e+03	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003592	F	Potassium - ICP-Fusion	ug/g	94.60	<5.00e-01	5.88e+06	n/a	n/a	n/a	n/a	1.05e+03	n/a
S961003592	F	Lanthanum - ICP-Fusion	ug/g	100.2	<5.00e-02	<1.04e+03	<1.06e3	n/a	n/a	97.50	1.05e+03	n/a
S961003592	F	Lithium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	95.60	209.0	n/a
S961003592	F	Magnesium - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	93.10	2.09e+03	n/a
S961003592	F	Manganese - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	93.60	209.0	n/a
S961003592	F	Molybdenum - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.04e+03	<1.06e3	n/a	n/a	97.80	1.05e+03	n/a
S961003592	F	Sodium - ICP-Fusion	ug/g	96.20	1.320	2.39e+05	2.32e+05	2.36e+05	2.97	111.0	2.09e+03	n/a
S961003592	F	Neodymium - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	97.70	2.09e+03	n/a
S961003592	F	Nickel - ICP-Fusion	ug/g	99.60	2.38e-01	1.09e+03	1.82e+03	1.46e+03	50.2	96.80	4.18e+03	n/a
S961003592	F	Phosphorus - ICP-Fusion	ug/g	101.4	<2.00e-01	<4.18e+03	<4.25e3	n/a	n/a	102.0	4.18e+03	n/a
S961003592	F	Lead - ICP-Fusion	ug/g	100.2	<1.00e-01	<3.42e+03	<3.69e+03	n/a	n/a	96.70	2.09e+03	n/a
S961003592	F	Sulfur - ICP-Fusion	ug/g	98.00	<1.00e-01	<3.42e+03	<3.69e+03	n/a	n/a	96.70	2.09e+03	n/a
S961003592	F	Antimony - ICP-Fusion	ug/g	94.60	<6.00e-02	<1.25e+03	<1.27e3	n/a	n/a	92.50	1.25e+03	n/a
S961003592	F	Selenium - ICP-Fusion	ug/g	98.80	<5.00e-02	<2.09e+03	<2.12e3	n/a	n/a	95.30	2.09e+03	n/a
S961003592	F	Silicon - ICP-Fusion	ug/g	96.60	<1.00e-01	<2.09e+03	<2.12e3	n/a	n/a	94.30	2.09e+03	n/a
S961003592	F	Samarium - ICP-Fusion	ug/g	97.40	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	94.30	2.09e+03	n/a
S961003592	F	Strontium - ICP-Fusion	ug/g	96.40	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	94.00	209.0	n/a
S961003592	F	Titanium - ICP-Fusion	ug/g	97.80	<2.00e-01	<4.18e+03	<4.25e3	n/a	n/a	94.00	209.0	n/a
S961003592	F	Thallium - ICP-Fusion	ug/g	94.60	<5.00e-01	<1.04e+04	<1.06e4	n/a	n/a	89.00	1.05e+04	n/a
S961003592	F	Vanadium - ICP-Fusion	ug/g	99.40	<5.00e-02	6.57e+02	<1.06e3	n/a	n/a	97.10	1.05e+04	n/a
S961003592	F	Zinc - ICP-Fusion	ug/g	101.0	<1.00e-02	6.57e+02	770.0	713.5	15.8	99.10	209.0	n/a
S961003592	F	Zirconium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.09e+02	<2.12e2	n/a	n/a	94.90	209.0	n/a
S961003592	F	Alpha of digested solid	ug/g	110.2	<1.80e-03	1.40e-02	1.14e-02	1.27e-02	20.5	n/a	5.00e-03	3.65e+01
S961003592	A	Silver - ICP-Acid Digest	ug/g	24.80	<1.00e-02	1.70e-02	17.70	17.45	2.87	n/a	3.960	n/a
S961003592	A	Aluminum - ICP-Acid Digest	ug/g	98.00	<1.44e-01	1.94e+04	1.96e+04	1.98e+04	1.03	n/a	39.60	n/a
S961003592	A	Arsenic - ICP-Acid Digest	ug/g	121.2	4.34e-01	60.70	63.00	61.85	3.72	n/a	19.80	n/a
S961003592	A	Boron - ICP-Acid Digest	ug/g	98.60	<5.00e-02	<1.980	<2.07e1	n/a	n/a	n/a	19.80	n/a
S961003592	A	Beryllium - ICP-Acid Digest	ug/g	97.40	<5.00e-03	<1.980	<2.07e0	n/a	n/a	n/a	1.980	n/a
S961003592	A	Bismuth - ICP-Acid Digest	ug/g	95.20	<1.00e-01	<39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a
S961003592	A	Calcium - ICP-Acid Digest	ug/g	100.2	2.320	3.10e+02	532.0	421.0	32.7	n/a	39.60	n/a
S961003592	A	Cadmium - ICP-Acid Digest	ug/g	93.80	<5.00e-03	8.520	8.410	8.465	1.30	n/a	1.980	n/a
S961003592	A	Cerium - ICP-Acid Digest	ug/g	100.6	<1.00e-01	<39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a
S961003592	A	Cobalt - ICP-Acid Digest	ug/g	96.80	<2.00e-02	<7.920	<8.29e0	n/a	n/a	n/a	7.920	n/a
S961003592	A	Chromium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	1.99e+03	2.08e+03	2.04e+03	4.42	n/a	3.960	n/a
S961003592	A	Copper - ICP-Acid Digest	ug/g	95.40	<1.00e-02	3.280	<4.14e0	n/a	n/a	n/a	3.960	n/a
S961003592	A	Iron - ICP-Acid Digest	ug/g	96.40	7.70e-02	2.91e+02	123.0	140.0	24.3	n/a	19.80	n/a
S961003592	A	Potassium - ICP-Acid Digest	ug/g	98.40	<5.00e-01	1.97e+02	3.08e+03	3.00e+03	5.68	n/a	19.80	n/a
S961003592	A	Lanthanum - ICP-Acid Digest	ug/g	100.4	<5.00e-02	<19.80	<2.07e1	n/a	n/a	n/a	19.80	n/a
S961003592	A	Lithium - ICP-Acid Digest	ug/g	99.40	<1.00e-02	82.50	6.750	44.62	170	n/a	3.960	n/a
S961003592	A	Magnesium - ICP-Acid Digest	ug/g	91.00	<1.00e-01	<39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a
S961003592	A	Manganese - ICP-Acid Digest	ug/g	93.00	<1.00e-02	33.20	34.30	33.75	3.26	n/a	3.960	n/a
S961003592	A	Molybdenum - ICP-Acid Digest	ug/g	96.40	<5.00e-02	48.40	49.80	49.10	2.85	n/a	19.80	n/a
S961003592	A	Sodium - ICP-Acid Digest	ug/g	138.6	1.110	1.93e+05	1.97e+05	1.98e+05	2.05	n/a	39.60	n/a
S961003592	A	Neodymium - ICP-Acid Digest	ug/g	99.60	<1.00e-01	<39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a
S961003592	A	Nickel - ICP-Acid Digest	ug/g	96.40	<2.00e-02	29.80	27.80	28.80	6.94	n/a	7.920	n/a
S961003592	A	Phosphorus - ICP-Acid Digest	ug/g	94.60	<2.00e-01	2.52e+03	2.58e+03	2.53e+03	2.35	n/a	79.20	n/a
S961003592	A	Lead - ICP-Acid Digest	ug/g	90.80	<1.00e-01	43.20	41.70	43.60	8.72	n/a	39.60	n/a
S961003592	A	Sulfur - ICP-Acid Digest	ug/g	93.20	<1.00e-01	2.61e+03	2.70e+03	2.68e+03	3.39	n/a	39.60	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
5961003905	A	Antimony -ICP-Acid Digest	ug/g	89.40	<6.00e-02	< 23.80	<2.49e1	n/a	n/a	n/a	23.80	n/a	n/a
5961003905	A	Selenium -ICP-Acid Digest	ug/g	89.20	<1.00e-01	< 39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a	n/a
5961003905	A	Silicon -ICP-Acid Digest	ug/g	264.0	2.20e-01	1.15e+02	102.0	108.5	12.0	n/a	19.80	n/a	n/a
5961003905	A	Samarium -ICP-Acid Digest	ug/g	97.20	<1.00e-01	< 39.60	<4.14e1	n/a	n/a	n/a	39.60	n/a	n/a
5961003905	A	Strontium -ICP-Acid Digest	ug/g	97.20	<1.00e-02	< 3.960	<4.14e0	n/a	n/a	n/a	3.960	n/a	n/a
5961003905	A	Titanium -ICP-Acid Digest	ug/g	89.80	<2.00e-02	< 3.960	<4.14e0	n/a	n/a	n/a	3.960	n/a	n/a
5961003905	A	Thallium -ICP-Acid Digest	ug/g	89.80	<2.00e-01	< 79.20	<8.29e1	n/a	n/a	n/a	79.20	n/a	n/a
5961003905	A	Uranium -ICP-Acid Digest	ug/g	95.40	<5.00e-01	<1.98e+02	<2.07e2	n/a	n/a	n/a	198.0	n/a	n/a
5961003905	A	Vanadium -ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 19.80	<2.07e1	n/a	n/a	n/a	19.80	n/a	n/a
5961003905	A	Zinc -ICP-Acid Digest	ug/g	90.20	3.90e-02	36.50	37.90	37.20	3.76	n/a	3.960	n/a	n/a
5961003905	A	Zirconium -ICP-Acid Digest	ug/g	96.40	<1.00e-02	25.20	25.30	25.25	0.40	n/a	3.960	n/a	n/a
5961003914	M	Acetate by IC-DIONE 4000/4500	ug/g	105.6	<2.00e-02	6.55e+02	645.0	650.0	1.54	n/a	48.10	n/a	n/a
5961003914	M	Formate by IC-DIONE 4000/4500	ug/g	95.83	<2.10e-02	1.10e+03	1.02e+03	1.09e+03	7.55	n/a	50.50	n/a	n/a
5961003914	M	Fluoride-IC-DIONE 4000/4500	ug/g	102.5	<1.30e-02	1.21e+03	1.12e+03	1.17e+03	7.23	n/a	59.48	n/a	n/a
5961003914	M	Chloride-IC-DIONE 4000/4500	ug/g	105.2	2.50e-02	4.44e+03	4.43e+03	4.43e+03	0.23	n/a	78.06	n/a	n/a
5961003914	M	Nitrite-IC - DIONE 4000/4500	ug/g	102.9	<1.07e-01	5.31e+04	5.28e+04	5.30e+04	0.57	n/a	491.2	n/a	n/a
5961003914	M	Bromide by Ion Chromatograph	ug/g	103.4	<1.26e-01	<5.39e2	<5.39e2	n/a	n/a	n/a	578.6	n/a	n/a
5961003914	M	Nitrate by IC-DIONE 4000/4500	ug/g	104.0	1.68e-01	1.65e+05	1.69e+05	1.67e+05	2.40	n/a	642.4	n/a	n/a
5961003914	M	Phosphate-IC-DIONE 4000/4500	ug/g	104.0	<1.19e-01	6.70e+03	5.85e+03	6.27e+03	13.5	n/a	525.8	n/a	n/a
5961003914	M	Sulfate by IC-DIONE 4000/4500	ug/g	104.6	<1.36e-01	6.44e+03	6.11e+03	6.27e+03	5.26	n/a	623.9	n/a	n/a
5961003914	M	Oxalate by IC-DIONE 4000/4500	ug/g	102.8	<1.05e-01	4.42e+03	4.35e+03	4.39e+03	1.60	n/a	482.2	n/a	n/a

CORE NUMBER: 152
SEGMENT #: Core Composite #22 ⁴⁵ 11/9/97

SEGMENT PORTION: Core Composite

Table 1. Data Summary Table
AN-105

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
595104208		Bulk Density of Sample	g/ml	n/a	n/a	1.570	n/a	n/a	n/a	n/a	5.00e-01	n/a

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

Table 1. Data Summary Table
AN-105

CORE NUMBER: 153

SEGMENT #: 1

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	% Spk Rec %	Det Limit	Count	Err%
S961004433	D	DSC Exotherm Dry, Calculated	Joules/g Dry	n/a	n/a	25.20	31.50	28.35	22.2	n/a	n/a	n/a	n/a
S961004433	D	DSC Exotherm on Perkin Elmer	Joules/g	98.49	n/a	12.70	15.90	14.30	22.4	n/a	n/a	n/a	n/a
S961004433	D	% Water by Gravimetry	Sp.G.	100.7	n/a	1.431	1.440	1.435	0.63	n/a	1.00e-03	n/a	n/a
S961004433	D	% Water by Total on Perkin Elmer	Sp.G.	99.11	n/a	49.71	49.30	49.50	0.83	n/a	n/a	n/a	n/a
S961004433	D	Aluminum-ICP-Acid Dil.	ug/ml	99.60	<1.00e-02	20.00	18.40	4.50e+04	7.33	84.70	4.010	n/a	n/a
S961004433	D	Aluminum-ICP-Acid Dil.	ug/ml	98.80	<1.00e-02	4.67e+04	4.34e+04	4.50e+04	7.33	95.40	40.10	n/a	n/a
S961004433	D	Aluminum-ICP-Acid Dil.	ug/ml	104.6	<1.00e-01	<40.10	<4.01e1	74.60	2.41	102.0	20.10	n/a	n/a
S961004433	D	Boron-ICP-Acid Dil.	ug/ml	101.6	<1.00e-02	75.50	73.70	74.60	2.41	102.0	20.10	n/a	n/a
S961004433	D	Boron-ICP-Acid Dil.	ug/ml	101.6	<1.00e-02	<20.10	<2.01e1	n/a	n/a	105.0	2.000	n/a	n/a
S961004433	D	Beryllium-ICP-Acid Dil.	ug/ml	104.6	<1.00e-02	<2.000	<2.00e0	n/a	n/a	102.0	40.10	n/a	n/a
S961004433	D	Bismuth-ICP-Acid Dil.	ug/ml	101.4	<1.00e-01	<40.10	<4.01e1	n/a	n/a	105.0	40.10	n/a	n/a
S961004433	D	Calcium-ICP-Acid Dil.	ug/ml	99.80	<1.00e-01	<40.10	<4.01e1	n/a	n/a	105.0	2.000	n/a	n/a
S961004433	D	Cadmium-ICP-Acid Dil.	ug/ml	100.4	<1.00e-03	<2.000	<2.00e0	n/a	n/a	102.0	40.10	n/a	n/a
S961004433	D	Cerium-ICP-Acid Dil.	ug/ml	100.4	<1.00e-01	<40.10	<4.01e1	n/a	n/a	105.0	4.010	n/a	n/a
S961004433	D	Cobalt-ICP-Acid Dil.	ug/ml	102.4	<1.00e-02	8.020	<8.02e0	383.5	6.00	79.50	4.010	n/a	n/a
S961004433	D	Chromium-ICP-Acid Dil.	ug/ml	105.6	<1.00e-02	3.95e+02	3.72e0	7.785	17.9	104.0	4.010	n/a	n/a
S961004433	D	Copper-ICP-Acid Dil.	ug/ml	101.2	<1.00e-02	8.480	7.090	7.785	17.9	104.0	4.010	n/a	n/a
S961004433	D	Iron-ICP-Acid Dil.	ug/ml	101.2	<1.00e-02	<20.10	<2.01e1	n/a	n/a	105.0	20.10	n/a	n/a
S961004433	D	Potassium-ICP-Acid Dil.	ug/ml	95.80	<1.00e-01	6.92e+03	6.40e+03	6.66e+03	7.81	116.0	200.0	n/a	n/a
S961004433	D	Lanthanum-ICP-Acid Dil.	ug/ml	102.2	<1.00e-02	<20.10	<2.01e1	n/a	n/a	104.0	20.10	n/a	n/a
S961004433	D	Lithium-ICP-Acid Dil.	ug/ml	102.4	<1.00e-02	<4.010	<4.01e0	n/a	n/a	110.0	4.010	n/a	n/a
S961004433	D	Magnesium-ICP-Acid Dil.	ug/ml	98.00	<1.00e-01	<40.10	<4.01e1	n/a	n/a	96.50	40.10	n/a	n/a
S961004433	D	Manganese-ICP-Acid Dil.	ug/ml	98.20	<1.00e-02	<4.010	<4.01e0	n/a	n/a	96.10	4.010	n/a	n/a
S961004433	D	Molybdenum-ICP-Acid Dil.	ug/ml	102.6	<1.00e-02	1.05e+02	100.0	102.5	4.88	104.0	20.10	n/a	n/a
S961004433	D	Sodium-ICP-Acid Dil.	ug/ml	99.20	<1.00e-01	2.69e+05	2.51e+05	2.60e+05	6.92	477.0	40.10	n/a	n/a
S961004433	D	Niobium-ICP-Acid Dil.	ug/ml	102.6	<1.00e-01	<40.10	<4.01e1	n/a	n/a	104.0	40.10	n/a	n/a
S961004433	D	Nickel-ICP-Acid Dil.	ug/ml	100.8	<1.00e-02	8.020	<8.02e0	n/a	n/a	100.0	8.020	n/a	n/a
S961004433	D	Phosphorus-ICP-Acid Dil.	ug/ml	104.6	<1.00e-01	6.56e+02	655.0	655.5	0.15	85.30	80.20	n/a	n/a
S961004433	D	Lead-ICP-Acid Dil.	ug/ml	100.4	<1.00e-01	50.60	46.40	48.50	8.66	100.0	40.10	n/a	n/a
S961004433	D	Sulfur-ICP-Acid Dil.	ug/ml	101.8	<1.00e-01	1.11e+03	1.06e+03	1.08e+03	6.51	103.0	40.10	n/a	n/a
S961004433	D	Selenium-ICP-Acid Dil.	ug/ml	97.40	<1.00e-02	3.05e+02	279.0	292.0	8.90	93.80	20.10	n/a	n/a
S961004433	D	Silicon-ICP-Acid Dil.	ug/ml	96.00	<1.00e-01	<24.10	<2.41e1	n/a	n/a	107.0	40.10	n/a	n/a
S961004433	D	Silicon-ICP-Acid Dil.	ug/ml	96.00	<1.00e-02	3.05e+02	279.0	292.0	8.90	93.80	20.10	n/a	n/a
S961004433	D	Silicon-ICP-Acid Dil.	ug/ml	98.80	<1.00e-01	<40.10	<4.01e1	n/a	n/a	98.50	40.10	n/a	n/a
S961004433	D	Strontium-ICP-Acid Dil.	ug/ml	100.0	<1.00e-02	<4.010	<4.01e0	n/a	n/a	101.0	4.010	n/a	n/a
S961004433	D	Titanium-ICP-Acid Dil.	ug/ml	99.40	<1.00e-02	<4.010	<4.01e0	n/a	n/a	99.90	4.010	n/a	n/a
S961004433	D	Thallium-ICP-Acid Dil.	ug/ml	97.20	<1.00e-01	<80.20	<8.02e1	n/a	n/a	91.80	80.20	n/a	n/a
S961004433	D	Uranium-ICP-Acid Dil.	ug/ml	98.00	<1.00e-01	<2.00e+02	<2.00e2	n/a	n/a	98.00	200.0	n/a	n/a
S961004433	D	Vanadium-ICP-Acid Dil.	ug/ml	102.2	<1.00e-02	<20.10	<2.01e1	n/a	n/a	102.0	20.10	n/a	n/a
S961004433	D	Zinc-ICP-Acid Dil.	ug/ml	102.6	<1.00e-02	81.10	54.10	67.60	39.9	98.80	4.010	n/a	n/a
S961004433	D	Zirconium-ICP-Acid Dil.	ug/ml	100.2	<1.00e-02	4.070	<4.07e0	n/a	n/a	102.0	4.010	n/a	n/a
S961004433	D	Acetate by IC-Dioxen 4000/4500	ug/ml	100.8	<1.00e-02	1.52e+03	694.0	1.11e+03	74.6	n/a	103.0	n/a	n/a
S961004433	D	Formate by Ion Chromatograph	ug/ml	104.2	<2.10e-02	1.84e+03	1.86e+03	1.85e+03	1.08	n/a	106.2	n/a	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R. A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec. %	Det Limit	Count Err%
S96T004433		Fluoride-IC-Dionex 4000/4500	ug/mL	102.7	<1.20e-02	4.61e+02	456.0	458.4	1.09	70.51	49.69	n/a
S96T004433		Chloride-IC-Dionex 4000/4500	ug/mL	106.2	<1.70e-02	1.06e+04	1.04e+04	1.05e+04	1.90	92.15	70.40	n/a
S96T004433		Nitrite-IC - Dionex 4000/4500	ug/mL	102.4	<1.08e-01	1.29e+05	1.27e+05	1.28e+05	1.56	92.45	447.2	n/a
S96T004433		Bromide by Ion Chromatograph	ug/mL	98.64	<1.25e-01	-5.18e+02	<5.18e+02	n/a	n/a	95.08	517.6	n/a
S96T004433		Nitrate by IC-Dionex 4000/4500	ug/mL	103.2	<1.39e-01	1.66e+05	1.64e+05	1.65e+05	1.21	86.70	575.6	n/a
S96T004433		Phosphate-IC-Dionex 4000/4500	ug/mL	100.4	<1.20e-01	6.59e+02	724.0	689.6	10.0	100.6	496.9	n/a
S96T004433		Sulfate by IC-Dionex 4000/4500	ug/mL	102.1	<1.38e-01	1.10e+03	1.07e+03	1.09e+03	2.76	101.1	571.5	n/a
S96T004433		Oxalate by IC-Dionex 4000/4500	ug/mL	99.62	<1.05e-01	-4.39e+02	474.0	n/a	n/a	100.9	434.8	n/a
S96T004433		Alpha in Liquid Samples	ug/mL	91.41	<2.20e-02	-2.53e-02	<2.84E-2	n/a	n/a	n/a	3.80e-02	5.00E+02

CORE NUMBER : 153
SEGMENT #: 2

SEGMENT PORTION: Drainable Liquid

Table 1. Data Summary Table
AH-105

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004308	D	DSC Exotherm using Mettler	Joules/g	94.55	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T004308	D	DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T004308	D	Specific Gravity	Sp.G.	97.00	n/a	1.534	1.454	1.494	5.35	n/a	1.00e-02	n/a
S96T004308	D	% Water by TGA using Mettler	%	98.72	n/a	50.38	50.23	50.30	0.30	n/a	n/a	n/a
S96T004308	D	Aluminum-ICP-Acid Dill.	ug/mL	100.0	<1.00e-02	4.39e+04	17.00	6.82	92.60	n/a	6.010	n/a
S96T004308	D	Silver-ICP-Acid Dill.	ug/mL	98.00	<5.00e-02	75.00	<6.010	74.70	96.80	n/a	30.10	n/a
S96T004308	D	Aluminum-ICP-Acid Dill.	ug/mL	104.4	<1.00e-01	3.010	<3.010	3.010	103.0	n/a	60.10	n/a
S96T004308	D	Barium-ICP-Acid Dill.	ug/mL	99.80	<5.00e-02	3.000	<3.000	3.000	98.40	n/a	30.10	n/a
S96T004308	D	Bismuth-ICP-Acid Dill.	ug/mL	101.8	<1.00e-01	6.010	<6.010	6.010	102.0	n/a	60.10	n/a
S96T004308	D	Calcium-ICP-Acid Dill.	ug/mL	101.2	<5.00e-03	3.000	<3.000	3.000	98.90	n/a	3.000	n/a
S96T004308	D	Cadmium-ICP-Acid Dill.	ug/mL	99.00	<1.00e-01	6.010	<6.010	6.010	98.00	n/a	60.10	n/a
S96T004308	D	Cobalt-ICP-Acid Dill.	ug/mL	102.2	<2.00e-02	12.00	<12.00	12.00	99.80	n/a	12.00	n/a
S96T004308	D	Chromium-ICP-Acid Dill.	ug/mL	102.2	<1.00e-02	2.22e+02	215.0	218.5	3.20	100.0	6.010	n/a
S96T004308	D	Copper-ICP-Acid Dill.	ug/mL	97.00	<1.00e-02	6.010	<6.010	6.010	94.90	n/a	6.010	n/a
S96T004308	D	Iron-ICP-Acid Dill.	ug/mL	101.6	<5.00e-02	3.010	<3.010	3.010	99.90	n/a	30.10	n/a
S96T004308	D	Potassium-ICP-Acid Dill.	ug/mL	98.80	<5.00e-01	6.73e+03	6.57e+03	6.65e+03	2.41	102.0	300.0	n/a
S96T004308	D	Lanthanum-ICP-Acid Dill.	ug/mL	100.4	<5.00e-02	3.010	<3.010	3.010	99.70	n/a	30.10	n/a
S96T004308	D	Lithium-ICP-Acid Dill.	ug/mL	101.2	<1.00e-02	6.010	<6.010	6.010	104.0	n/a	60.10	n/a
S96T004308	D	Magnesium-ICP-Acid Dill.	ug/mL	97.60	<1.00e-01	6.010	<6.010	6.010	95.60	n/a	60.10	n/a
S96T004308	D	Manganese-ICP-Acid Dill.	ug/mL	98.80	<1.00e-02	6.010	<6.010	6.010	94.90	n/a	6.010	n/a
S96T004308	D	Nickel-ICP-Acid Dill.	ug/mL	102.0	<5.00e-02	1.01e+02	97.20	99.10	3.83	101.0	30.10	n/a
S96T004308	D	Sodium-ICP-Acid Dill.	ug/mL	98.00	<1.00e-01	2.58e+05	2.48e+05	2.53e+05	3.95	-2.29e+02	60.10	n/a
S96T004308	D	Neodymium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	6.010	<6.010	6.010	99.20	n/a	60.10	n/a
S96T004308	D	Nickel-ICP-Acid Dill.	ug/mL	101.2	<2.00e-02	12.00	<12.00	12.00	98.90	n/a	12.00	n/a
S96T004308	D	Phosphorus-ICP-Acid Dill.	ug/mL	101.4	<5.00e-01	3.94e+02	368.0	381.0	6.82	101.0	120.0	n/a
S96T004308	D	Lead-ICP-Acid Dill.	ug/mL	100.8	<1.00e-01	6.010	<6.010	6.010	106.0	n/a	60.10	n/a
S96T004308	D	Sulfur-ICP-Acid Dill.	ug/mL	97.20	<5.00e-02	36.10	<36.10	36.10	96.80	n/a	36.10	n/a
S96T004308	D	Antimony-ICP-Acid Dill.	ug/mL	98.40	<1.00e-01	6.010	<6.010	6.010	107.0	n/a	60.10	n/a
S96T004308	D	Selenium-ICP-Acid Dill.	ug/mL	94.80	<5.00e-02	1.87e+02	181.0	184.0	3.26	96.30	30.10	n/a
S96T004308	D	Silicon-ICP-Acid Dill.	ug/mL	97.20	<1.00e-01	6.010	<6.010	6.010	96.40	n/a	60.10	n/a
S96T004308	D	Strontium-ICP-Acid Dill.	ug/mL	98.60	<1.00e-02	6.010	<6.010	6.010	96.90	n/a	6.010	n/a
S96T004308	D	Titanium-ICP-Acid Dill.	ug/mL	98.40	<1.00e-02	<1.20e+02	<3.00e+02	<3.00e+02	94.60	n/a	120.0	n/a
S96T004308	D	Thallium-ICP-Acid Dill.	ug/mL	95.90	<5.00e-01	<3.00e+02	<3.00e+02	<3.00e+02	97.00	n/a	300.0	n/a
S96T004308	D	Uranium-ICP-Acid Dill.	ug/mL	101.6	<5.00e-02	23.50	<23.50	22.30	10.8	99.70	6.010	n/a
S96T004308	D	Vanadium-ICP-Acid Dill.	ug/mL	102.4	<1.00e-02	6.010	<6.010	6.010	98.60	n/a	6.010	n/a
S96T004308	D	Zinc-ICP-Acid Dill.	ug/mL	99.00	<2.00e-02	5.80e+02	650.0	615.0	11.4	75.08	82.82	n/a
S96T004308	D	Zirconium-ICP-Acid Dill.	ug/mL	96.77	<2.00e-02	1.69e+03	1.81e+03	1.75e+03	6.86	102.5	86.96	n/a
S96T004308	D	Formate by Ion Chromatograph	ug/mL	98.33	<2.10e-02							

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004308		Fluoride-IC-Dionex 4000/4500	ug/mL	99.52	<1.30e-02	4.50e+02	431.0	440.2	4.31	75.25	53.83	n/a
S96T004308		Chloride-IC-Dionex 4000/4500	ug/mL	103.0	<1.70e-02	1.02e+04	1.00e+04	1.01e+04	1.98	95.95	70.40	n/a
S96T004308		Nitrite-IC - Dionex 4000/4500	ug/mL	101.1	<1.07e-01	1.25e+05	1.23e+05	1.24e+05	1.61	90.42	443.1	n/a
S96T004308		Bromide by Ion Chromatograph	ug/mL	99.49	<1.26e-01	8.44e+02	857.0	850.7	1.53	93.21	521.8	n/a
S96T004308		Nitrate by IC-Dionex 4000/4500	ug/mL	104.4	1.50e-01	1.62e+05	1.62e+05	1.62e+05	0.00	91.25	579.7	n/a
S96T004308		Phosphate-IC-Dionex 4000/4500	ug/mL	101.3	<1.19e-01	6.34e+02	606.0	620.0	4.52	100.9	492.4	n/a
S96T004308		Sulfate by IC-Dionex 4000/4500	ug/mL	103.3	<1.36e-01	1.05e+03	1.17e+03	1.11e+03	10.8	101.1	562.8	n/a
S96T004308		Oxalate by IC-Dionex 4000/4500	ug/mL	101.3	<1.05e-01	<4.35e+02	<4.35e+02	n/a	n/a	101.3	434.8	n/a
S96T004308		Alpha in Liquid Samples	UCI/mL	95.31	1.10e-02	<1.36e-02	<7.10E-3	n/a	n/a	87.74	1.80e-02	5.00E+02

CORE NUMBER: 153
SEGMENT #: 3

SEGMENT PORTION: Drainable Liquid

Table 1. Data Summary Table
AN-105

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004434	D	DSC Exotherm using Mettler	Joules/g	106.5	n/a	16.80	19.90	18.35	16.9	n/a	n/a	n/a
S96T004434	D	DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	34.50	40.80	37.65	16.7	n/a	n/a	n/a
S96T004434	D	Specific Gravity	Sp.G.	100.7	n/a	1.459	1.458	1.458	0.07	n/a	1.00e-03	n/a
S96T004434	D	% Water by IGA using Mettler	%	98.29	n/a	51.07	51.41	51.24	0.66	n/a	n/a	n/a
S96T004434	D	Aluminum-ICP-Acid Dill.	ug/mL	99.60	<1.00e-02	17.80	18.70	18.25	4.93	90.40	4.010	n/a
S96T004434	D	Aluminum-ICP-Acid Dill.	ug/mL	98.80	<5.00e-02	3.94e+04	4.22e+04	4.08e+04	6.86	1.50e+03	20.10	n/a
S96T004434	D	Arsenic-ICP-Acid Dill.	ug/mL	104.6	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	89.90	40.10	n/a
S96T004434	D	Boron-ICP-Acid Dill.	ug/mL	101.6	<5.00e-02	65.40	76.40	70.90	15.5	104.0	20.10	n/a
S96T004434	D	Beryllium-ICP-Acid Dill.	ug/mL	100.2	<2.00e-02	< 2.000	<2.00e0	n/a	n/a	104.0	20.10	n/a
S96T004434	D	Bismuth-ICP-Acid Dill.	ug/mL	104.6	<5.00e-03	< 40.10	<4.01e1	n/a	n/a	101.0	40.10	n/a
S96T004434	D	Calcium-ICP-Acid Dill.	ug/mL	101.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	107.0	40.10	n/a
S96T004434	D	Cadmium-ICP-Acid Dill.	ug/mL	99.80	<1.00e-03	< 2.000	<2.00e0	n/a	n/a	101.0	2.000	n/a
S96T004434	D	Cerium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	102.0	40.10	n/a
S96T004434	D	Cobalt-ICP-Acid Dill.	ug/mL	102.4	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	102.0	8.020	n/a
S96T004434	D	Chromium-ICP-Acid Dill.	ug/mL	101.4	<1.00e-02	2.78e+02	276.0	277.0	0.72	95.70	4.010	n/a
S96T004434	D	Copper-ICP-Acid Dill.	ug/mL	105.6	<1.00e-02	10.20	6.740	8.470	40.9	103.0	4.010	n/a
S96T004434	D	Iron-ICP-Acid Dill.	ug/mL	101.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	103.0	20.10	n/a
S96T004434	D	Potassium-ICP-Acid Dill.	ug/mL	95.80	<5.00e-01	6.07e+03	6.60e+03	6.34e+03	8.37	327.0	200.10	n/a
S96T004434	D	Lanthanum-ICP-Acid Dill.	ug/mL	102.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	104.0	20.10	n/a
S96T004434	D	Lithium-ICP-Acid Dill.	ug/mL	102.4	<1.00e-02	13.80	11.00	12.40	22.6	105.0	4.010	n/a
S96T004434	D	Magnesium-ICP-Acid Dill.	ug/mL	98.00	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	96.30	40.10	n/a
S96T004434	D	Manganese-ICP-Acid Dill.	ug/mL	98.20	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	96.40	4.010	n/a
S96T004434	D	Molybdenum-ICP-Acid Dill.	ug/mL	102.6	<5.00e-02	92.30	103.0	97.65	11.0	108.0	20.10	n/a
S96T004434	D	Sodium-ICP-Acid Dill.	ug/mL	99.20	<1.00e-01	2.33e+05	2.44e+05	2.44e+05	9.41	9.74e+03	40.10	n/a
S96T004434	D	Neodymium-ICP-Acid Dill.	ug/mL	102.6	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	104.0	40.10	n/a
S96T004434	D	Nickel-ICP-Acid Dill.	ug/mL	100.8	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	99.90	8.020	n/a
S96T004434	D	Phosphorus-ICP-Acid Dill.	ug/mL	104.6	<2.00e-01	4.30e+02	435.0	432.5	1.16	98.10	80.20	n/a
S96T004434	D	Lead-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	44.80	50.80	47.80	12.6	103.0	40.10	n/a
S96T004434	D	Sulfur-ICP-Acid Dill.	ug/mL	101.8	<1.00e-02	7.82e+02	829.0	805.5	5.83	129.0	40.10	n/a
S96T004434	D	Antimony-ICP-Acid Dill.	ug/mL	97.40	<5.00e-02	< 24.10	<2.41e1	n/a	n/a	100.0	24.10	n/a
S96T004434	D	Selenium-ICP-Acid Dill.	ug/mL	98.60	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	106.0	40.10	n/a
S96T004434	D	Strontium-ICP-Acid Dill.	ug/mL	96.00	<5.00e-02	1.80e+02	196.0	188.0	8.51	103.0	20.10	n/a
S96T004434	D	Titanium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	99.00	40.10	n/a
S96T004434	D	Tungsten-ICP-Acid Dill.	ug/mL	99.40	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	101.0	4.010	n/a
S96T004434	D	Vanadium-ICP-Acid Dill.	ug/mL	100.0	<2.00e-01	< 80.20	<8.02e1	n/a	n/a	99.80	80.20	n/a
S96T004434	D	Uranium-ICP-Acid Dill.	ug/mL	98.00	<5.00e-01	<2.00e+02	<2.00e2	n/a	n/a	99.50	200.10	n/a
S96T004434	D	Zinc-ICP-Acid Dill.	ug/mL	102.2	<1.00e-02	< 20.10	<2.01e1	n/a	n/a	104.0	20.10	n/a
S96T004434	D	210Pb-ICP-Acid Dill.	ug/mL	102.6	<1.00e-02	55.40	47.90	51.65	14.5	104.0	4.010	n/a
S96T004434	D	Zirconium-ICP-Acid Dill.	ug/mL	100.2	<2.00e-02	< 4.010	<4.01e0	n/a	n/a	103.0	4.010	n/a
S96T004434	D	Acetate by IC-Dioxex 4000/4500	ug/mL	105.6	<2.00e-02	4.52e+02	456.0	454.1	0.88	91.13	42.42	n/a
S96T004434	D	Formate by Ion Chromatograph	ug/mL	95.83	<2.10e-02	1.04e+03	1.02e+03	1.03e+03	1.94	102.5	44.54	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Error
S96T004434		Fluoride-IC-Dionex 4000/4500	ug/mL	102.7	<1.20e-02	4.14e+02	433.0	423.8	4.49	n/a	49.69	n/a
S96T004434		Chloride-IC-Dionex 4000/4500	ug/mL	106.2	<1.70e-02	1.02e+04	9.89e+03	1.00e+04	3.09	n/a	70.40	n/a
S96T004434		Nitrite-IC - Dionex 4000/4500	ug/mL	102.4	<1.08e-01	1.23e+05	1.22e+05	1.22e+05	0.82	n/a	447.2	n/a
S96T004434		Bromide by Ion Chromatograph	ug/mL	98.64	<1.25e-01	9.52e+02	952.0	952.1	0.00	n/a	517.6	n/a
S96T004434		Nitrate by IC-Dionex 4000/4500	ug/mL	103.2	<1.39e-01	1.61e+05	1.61e+05	1.61e+05	0.00	n/a	575.6	n/a
S96T004434		Phosphate-IC-Dionex 4000/4500	ug/mL	100.4	<1.20e-01	6.01e+02	745.0	673.0	21.4	n/a	496.9	n/a
S96T004434		Sulfate by IC-Dionex 4000/4500	ug/mL	102.1	<1.38e-01	1.10e+03	1.08e+03	1.09e+03	1.83	n/a	571.5	n/a
S96T004434		Oxalate by IC-Dionex 4000/4500	ug/mL	99.62	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	n/a	434.8	n/a
S96T004434		Alpha In Liquid Samples	uCi/mL	91.41	<2.20e-02	<1.54e-02	<1.87E-2	n/a	n/a	n/a	3.80e-02	5.00e+02

Table 1. Data Summary Table
AN-105

CORE NUMBER: 153
SEGMENT #: 5

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S96T004419	D	DSC Exotherm using Mettler	Joules/g	99.47	n/a	38.40	205.2	111.6	168	n/a	n/a	n/a	n/a
S96T004419	D	DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	16.40	414.0	225.2	n/a	n/a	n/a	n/a	n/a
S96T004419	1	DSC Exotherm DRY	Joules/g DRY	n/a	n/a	0.00e+00	29.30	14.65	200	n/a	n/a	n/a	n/a
S96T004419	1	DSC Exotherm on Perkin Elmer	Joules/g	96.34	n/a	0.00e+00	14.53	7.265	200	n/a	n/a	n/a	n/a
S96T004419		Specific Gravity	Sp.-G.	97.00	n/a	1.444	1.435	1.440	0.63	n/a	1.00e-02	n/a	n/a
S96T004419		% Water by TGA using Mettler	%	100.0	n/a	50.41	0.26	50.47	0.26	n/a	n/a	n/a	n/a
S96T004419	D	Silver-ICP-Acid Dil.	ug/mL	101.0	<1.00e-02	19.00	18.90	18.95	0.53	90.70	6.010	n/a	n/a
S96T004419	D	Aluminum-ICP-Acid Dil.	ug/mL	98.60	<5.00e-02	4.42e+04	4.37e+04	4.40e+04	1.14	180.0	30.10	n/a	n/a
S96T004419	D	Arsenic-ICP-Acid Dil.	ug/mL	104.0	<1.00e-01	< 6.010	<6.01e1	n/a	n/a	94.40	60.10	n/a	n/a
S96T004419	D	Boron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 30.10	<3.01e1	75.45	4.64	98.80	30.10	n/a	n/a
S96T004419	D	Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	98.80	30.10	n/a	n/a
S96T004419	D	Beryllium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	101.0	3.000	n/a	n/a
S96T004419	D	Bismuth-ICP-Acid Dil.	ug/mL	103.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	101.0	60.10	n/a	n/a
S96T004419	D	Calcium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a	n/a
S96T004419	D	Cadmium-ICP-Acid Dil.	ug/mL	100.4	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	101.0	3.000	n/a	n/a
S96T004419	D	Cerium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	100.0	60.10	n/a	n/a
S96T004419	D	Cobalt-ICP-Acid Dil.	ug/mL	103.4	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	102.0	12.00	n/a	n/a
S96T004419	D	Chromium-ICP-Acid Dil.	ug/mL	102.6	<1.00e-02	2.24e+02	219.0	221.5	2.26	101.0	6.010	n/a	n/a
S96T004419	D	Copper-ICP-Acid Dil.	ug/mL	104.8	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	103.0	6.010	n/a	n/a
S96T004419	D	Iron-ICP-Acid Dil.	ug/mL	101.8	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a	n/a
S96T004419	D	Potassium-ICP-Acid Dil.	ug/mL	101.8	<5.00e-01	6.96e+03	7.03e+03	7.00e+03	1.00	111.0	300.0	n/a	n/a
S96T004419	D	Lanthanum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a	n/a
S96T004419	D	Lithium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	104.0	6.010	n/a	n/a
S96T004419	D	Magnesium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	95.60	60.10	n/a	n/a
S96T004419	D	Manganese-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	96.50	6.010	n/a	n/a
S96T004419	D	Molybdenum-ICP-Acid Dil.	ug/mL	103.8	<5.00e-02	1.02e+02	103.0	102.5	0.98	104.0	30.10	n/a	n/a
S96T004419	D	Sodium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	2.62e+05	2.58e+05	2.60e+05	1.54	307.0	60.10	n/a	n/a
S96T004419	D	Neodymium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a	n/a
S96T004419	D	Nickel-ICP-Acid Dil.	ug/mL	102.4	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	101.0	12.00	n/a	n/a
S96T004419	D	Phosphorus-ICP-Acid Dil.	ug/mL	104.0	<2.00e-01	3.86e+02	359.0	372.5	7.25	102.0	120.0	n/a	n/a
S96T004419	D	Lead-ICP-Acid Dil.	ug/mL	102.4	<1.00e-01	< 60.10	<6.01e1	1.02e+03	0.00	109.0	60.10	n/a	n/a
S96T004419	D	Sulfur-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	1.02e+03	< 36.10	1.02e+03	0.00	106.0	60.10	n/a	n/a
S96T004419	D	Antimony-ICP-Acid Dil.	ug/mL	97.60	<5.00e-02	< 36.10	<3.61e1	n/a	n/a	99.50	36.10	n/a	n/a
S96T004419	D	Selenium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a	n/a
S96T004419	D	Silicon-ICP-Acid Dil.	ug/mL	97.20	<5.00e-02	2.32e+02	229.0	230.5	1.30	98.10	30.10	n/a	n/a
S96T004419	D	Samarium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	99.40	60.10	n/a	n/a
S96T004419	D	Samarium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	99.00	6.010	n/a	n/a
S96T004419	D	Strontium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	98.80	6.010	n/a	n/a
S96T004419	D	Titanium-ICP-Acid Dil.	ug/mL	97.80	<5.00e-01	<1.20e+02	<1.20e2	n/a	n/a	95.70	120.0	n/a	n/a
S96T004419	D	Thallium-ICP-Acid Dil.	ug/mL	97.80	<5.00e-01	<1.20e+02	<1.20e2	n/a	n/a	98.50	300.0	n/a	n/a
S96T004419	D	Uranium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a	n/a
S96T004419	D	Vanadium-ICP-Acid Dil.	ug/mL	104.4	<1.00e-02	41.80	41.70	41.75	0.24	104.0	6.010	n/a	n/a
S96T004419	D	Zinc-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	102.0	6.010	n/a	n/a
S96T004419	D	Zirconium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	102.0	6.010	n/a	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004419		Acetate by IC-Dionex 4000/4500	ug/mL	104.8	<2.00e-02	6.02e+02	533.0	567.8	12.2	81.45	103.0	n/a
S96T004419		Formate by Ion Chromatograph	ug/mL	104.2	<2.10e-02	1.63e+03	1.56e+03	1.60e+03	4.39	90.00	108.2	n/a
S96T004419		Fluoride-IC-Dionex 4000/4500	ug/mL	102.0	<1.30e-02	4.39e+02	447.0	442.9	1.81	n/a	53.83	n/a
S96T004419		Chloride-IC-Dionex 4000/4500	ug/mL	101.5	<1.70e-02	1.03e+04	1.00e+04	1.01e+04	2.96	n/a	70.40	n/a
S96T004419		Nitrite-IC - Dionex 4000/4500	ug/mL	99.45	<1.07e-01	1.24e+05	1.24e+05	1.24e+05	0.00	n/a	443.1	n/a
S96T004419		Bromide by Ion Chromatograph	ug/mL	100.8	<1.26e-01	1.04e+03	1.08e+03	1.06e+03	3.77	n/a	521.8	n/a
S96T004419		Nitrate by IC-Dionex 4000/4500	ug/mL	101.5	1.40e-01	1.70e+05	1.70e+05	1.70e+05	0.00	n/a	579.7	n/a
S96T004419		Phosphate-IC-Dionex 4000/4500	ug/mL	99.63	<1.19e-01	7.83e+02	636.0	709.3	20.7	n/a	492.4	n/a
S96T004419		Sulfate by IC-Dionex 4000/4500	ug/mL	101.6	<1.36e-01	1.08e+03	1.17e+03	1.12e+03	8.00	n/a	562.8	n/a
S96T004419		Oxalate by IC-Dionex 4000/4500	ug/mL	100.4	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	n/a	434.8	n/a
S96T004419		Alpha in Liquid Samples	uCi/mL	91.41	<2.20e-02	<1.54e-02	<1.87E-2	n/a	n/a	n/a	3.80e-02	5.00E+02

CORE NUMBER: 153
SEGMENT #: 6

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk %	Det Limit	Count Err%
S961004420	D	DSC Exotherm using Mettler	Joules/g	99.47	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961004420	D	DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961004420	D	Specific Gravity	Sp.G.	100.7	n/a	1.388	1.342	1.365	3.37	n/a	1.00e-03	n/a
S961004420	D	% Water by TGA using Mettler	%	100.0	<1.00e-02	50.10	51.00	50.55	1.78	n/a	6.010	n/a
S961004420	D	Silver-ICP-Acid Dil.	ug/mL	101.0	<1.00e-02	17.90	20.10	19.00	11.6	93.60	30.10	n/a
S961004420	D	Aluminum-ICP-Acid Dil.	ug/mL	98.60	<5.00e-02	4.13e+04	4.46e+04	4.30e+04	7.68	384.0	60.10	n/a
S961004420	D	Arsenic-ICP-Acid Dil.	ug/mL	104.0	<1.00e-01	<60.10	<6.01e1	n/a	n/a	94.90	30.10	n/a
S961004420	D	Boron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	<60.10	<6.01e1	72.85	10.3	99.70	30.10	n/a
S961004420	D	Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	<30.10	<3.01e1	n/a	n/a	98.80	30.10	n/a
S961004420	D	Beryllium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-03	<3.000	<3.00e0	n/a	n/a	102.0	3.000	n/a
S961004420	D	Bismuth-ICP-Acid Dil.	ug/mL	103.4	<1.00e-01	<60.10	<6.01e1	n/a	n/a	105.0	60.10	n/a
S961004420	D	Calcium-ICP-Acid Dil.	ug/mL	102.2	<5.00e-01	<60.10	<6.01e1	n/a	n/a	102.0	3.000	n/a
S961004420	D	Cadmium-ICP-Acid Dil.	ug/mL	100.4	<5.00e-03	<3.000	<3.00e0	n/a	n/a	101.0	60.10	n/a
S961004420	D	Cerium-ICP-Acid Dil.	ug/mL	103.4	<2.00e-02	<12.00	<1.20e1	n/a	n/a	103.0	12.00	n/a
S961004420	D	Cobalt-ICP-Acid Dil.	ug/mL	102.6	<1.00e-02	2.09e+02	225.0	217.0	7.37	111.0	6.010	n/a
S961004420	D	Chromium-ICP-Acid Dil.	ug/mL	104.8	<1.00e-02	<6.010	<6.01e0	n/a	n/a	104.0	6.010	n/a
S961004420	D	Copper-ICP-Acid Dil.	ug/mL	101.8	<5.00e-02	<30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a
S961004420	D	Potassium-ICP-Acid Dil.	ug/mL	101.8	<5.00e-01	6.64e+03	7.10e+03	6.87e+03	6.70	126.0	300.0	n/a
S961004420	D	Lanthanum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	<6.010	<6.01e0	n/a	n/a	102.0	30.10	n/a
S961004420	D	Lithium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	<6.010	<6.01e0	n/a	n/a	105.0	6.010	n/a
S961004420	D	Magnesium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-02	<60.10	<6.01e1	n/a	n/a	95.20	60.10	n/a
S961004420	D	Manganese-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	<6.010	<6.01e0	n/a	n/a	97.20	6.010	n/a
S961004420	D	Molybdenum-ICP-Acid Dil.	ug/mL	103.8	<5.00e-02	96.70	105.0	100.8	8.23	105.0	30.10	n/a
S961004420	D	Sodium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	2.46e+05	2.65e+05	2.56e+05	7.44	1.42e+03	60.10	n/a
S961004420	D	Neodymium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-01	<60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a
S961004420	D	Nickel-ICP-Acid Dil.	ug/mL	104.0	<2.00e-02	<12.00	<1.20e1	n/a	n/a	102.0	12.00	n/a
S961004420	D	Phosphorus-ICP-Acid Dil.	ug/mL	104.0	<2.00e-01	3.64e+02	370.0	367.0	1.63	109.0	120.0	n/a
S961004420	D	Lead-ICP-Acid Dil.	ug/mL	102.4	<1.00e-01	<60.10	<6.01e1	n/a	n/a	110.0	60.10	n/a
S961004420	D	Sulfur-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	9.39e+02	1.02e+03	979.5	8.27	117.0	60.10	n/a
S961004420	D	Antimony-ICP-Acid Dil.	ug/mL	97.60	<5.00e-02	<36.10	<3.61e1	n/a	n/a	99.80	36.10	n/a
S961004420	D	Selenium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	<60.10	<6.01e1	n/a	n/a	101.0	60.10	n/a
S961004420	D	Silicon-ICP-Acid Dil.	ug/mL	97.20	<5.00e-02	1.92e+02	206.0	199.0	7.04	102.0	30.10	n/a
S961004420	D	Samarium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	<60.10	<6.01e1	n/a	n/a	98.90	60.10	n/a
S961004420	D	Strontium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	<6.010	<6.01e0	n/a	n/a	99.10	6.010	n/a
S961004420	D	Titanium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	<6.010	<6.01e0	n/a	n/a	98.90	6.010	n/a
S961004420	D	Thallium-ICP-Acid Dil.	ug/mL	97.80	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	96.50	120.0	n/a
S961004420	D	Uranium-ICP-Acid Dil.	ug/mL	97.50	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	100.0	300.0	n/a
S961004420	D	Vanadium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	<30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a
S961004420	D	Zinc-ICP-Acid Dil.	ug/mL	104.4	<1.00e-02	<36.70	<3.67e1	37.95	6.59	105.0	6.010	n/a
S961004420	D	Zirconium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	<6.010	<6.01e0	n/a	n/a	102.0	6.010	n/a
S961004420	D	Acetate by IC-Bioxon 4000/4500	ug/mL	101.6	<2.00e-02	8.06e+02	951.0	878.4	16.5	n/a	103.0	n/a
S961004420	D	Formate by Ion Chromatograph	ug/mL	99.17	<2.10e-02	1.71e+03	2.04e+03	1.88e+03	17.6	n/a	108.2	n/a

13-Jan-1997 11:26:04
A-0002-3

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961004420		Fluoride-IC-Dionex 4000/4500	ug/mL	100.2	<1.30e-02	3.97e-02	4.15 0	4.06e-1	4.43	54.24	34.14		n/a
S961004420		Chloride-IC-Dionex 4000/4500	ug/mL	101.4	<1.70e-02	1.03e-04	1.05e+04	1.04e+04	1.92	99.62	44.64		n/a
S961004420		Nitrite-IC - Dionex 4000/4500	ug/mL	95.96	<1.08e-01	1.18e-03	1.16e+05	1.16e+05	0.00	97.41	44.7 2		n/a
S961004420		Bromide by Ion Chromatograph	ug/mL	96.64	<1.28e-01	1.18e-03	1.10e+03	1.13e+03	5.31	88.96	330.9		n/a
S961004420		Nitrate by IC-Dionex 4000/4500	ug/mL	99.66	1.59e-01	1.58e-03	1.58e+05	1.58e+05	0.00	95.45	575.6		n/a
S961004420		Phosphate-IC-Dionex 4000/4500	ug/mL	99.45	<1.19e-01	6.77e-02	665.0	671.3	1.64	98.72	312.2		n/a
S961004420		Sulfate by IC-Dionex 4000/4500	ug/mL	102.7	<1.36e-01	9.91e-02	296.0	948.6	8.96	100.5	356.9		n/a
S961004420		Oxalate by IC-Dionex 4000/4500	ug/mL	100.4	<1.05e-01	2.79e-02	296.0	288.6	6.59	99.05	275.7		n/a
S961004420		Alpha in Liquid Samples	ug/L/mL	91.41	<2.20e-02	<1.54e-02	<1.54e-2	n/a	n/a	n/a	3.80e-02	5.00E+02	

CORE NUMBER: 153
SEGMENT #: 7

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004421	D	DSC Exotherm Dry	Calculated	Joules/g	n/a	0.00e+00	69.00	34.50	200	n/a	n/a	n/a
S96T004421	D	DSC Exotherm on Perkin Elmer	Joules/g	98.49	n/a	0.00e+00	34.54	17.27	200	n/a	n/a	n/a
S96T004421	D	Specific Gravity	Sp.G.	100.7	n/a	1.511	1.503	1.507	0.53	n/a	1.00e-03	n/a
S96T004421	D	% Water by TGA on Perkin Elmer	%	99.11	n/a	49.93	49.97	49.95	0.08	n/a	n/a	n/a
S96T004421	D	Silver-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	18.20	19.60	18.90	7.41	86.80	4.010	n/a
S96T004421	D	Aluminum-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	4.29e+04	4.59e+04	4.69e+04	7.69	605.0	20.10	n/a
S96T004421	D	Arsenic-ICP-Acid Dil.	ug/mL	104.6	<1.00e-01	< 40.10	<4.01e1	< 40.10	n/a	91.70	40.10	n/a
S96T004421	D	Boron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	72.90	79.80	76.35	9.04	100.0	20.10	n/a
S96T004421	D	Barium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 20.10	<2.01e1	< 20.10	n/a	99.90	20.10	n/a
S96T004421	D	Beryllium-ICP-Acid Dil.	ug/mL	104.6	<5.00e-02	< 2.000	<2.00e0	n/a	n/a	104.0	2.000	n/a
S96T004421	D	Bismuth-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.50	40.10	n/a
S96T004421	D	Calcium-ICP-Acid Dil.	ug/mL	99.80	<1.00e-01	< 2.000	<2.00e0	n/a	n/a	108.0	40.10	n/a
S96T004421	D	Cadmium-ICP-Acid Dil.	ug/mL	100.4	<5.00e-03	< 40.10	<4.01e1	n/a	n/a	101.0	2.000	n/a
S96T004421	D	Cerium-ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	< 40.10	<4.01e1	n/a	n/a	101.0	40.10	n/a
S96T004421	D	Chromium-ICP-Acid Dil.	ug/mL	102.4	<2.00e-02	< 8.020	<8.02e0	323.0	39.0	120.0	8.020	n/a
S96T004421	D	Cobalt-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	< 20.10	<2.01e1	n/a	n/a	106.0	4.010	n/a
S96T004421	D	Copper-ICP-Acid Dil.	ug/mL	105.6	<5.00e-02	< 40.10	<4.01e1	n/a	n/a	104.0	4.010	n/a
S96T004421	D	Iron-ICP-Acid Dil.	ug/mL	101.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	104.0	20.10	n/a
S96T004421	D	Potassium-ICP-Acid Dil.	ug/mL	95.80	<5.00e-02	< 20.10	<2.01e1	6.71e+03	8.35	175.0	200.0	n/a
S96T004421	D	Lanthanum-ICP-Acid Dil.	ug/mL	102.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	103.0	20.10	n/a
S96T004421	D	Lithium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	< 21.00	<2.01e1	14.01	99.8	110.0	4.010	n/a
S96T004421	D	Magnesium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	< 4.010	<4.01e1	n/a	n/a	95.90	40.10	n/a
S96T004421	D	Manganese-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	< 4.010	<4.01e1	n/a	n/a	95.90	4.010	n/a
S96T004421	D	Molybdenum-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	98.10	103.0	100.5	4.87	105.0	20.10	n/a
S96T004421	D	Sodium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	2.55e+05	2.74e+05	2.64e+05	7.18	2.79e+03	40.10	n/a
S96T004421	D	Neodymium-ICP-Acid Dil.	ug/mL	102.6	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	103.0	40.10	n/a
S96T004421	D	Nickel-ICP-Acid Dil.	ug/mL	100.8	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	102.0	8.020	n/a
S96T004421	D	Phosphorus-ICP-Acid Dil.	ug/mL	104.6	<2.00e-01	4.94e+02	441.0	467.5	11.3	105.0	80.20	n/a
S96T004421	D	Lead-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	43.60	48.30	45.95	10.2	99.50	40.10	n/a
S96T004421	D	Sulfur-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	9.61e+02	1.07e+03	1.02e+03	10.7	118.0	40.10	n/a
S96T004421	D	Antimony-ICP-Acid Dil.	ug/mL	97.40	<5.00e-02	< 24.10	<2.41e1	n/a	n/a	101.0	24.10	n/a
S96T004421	D	Selenium-ICP-Acid Dil.	ug/mL	98.60	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	106.0	40.10	n/a
S96T004421	D	Silicon-ICP-Acid Dil.	ug/mL	96.00	<5.00e-02	2.25e+02	218.0	221.5	3.16	99.90	20.10	n/a
S96T004421	D	Strontium-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.50	40.10	n/a
S96T004421	D	Titanium-ICP-Acid Dil.	ug/mL	100.0	<1.00e-02	< 4.010	<4.01e1	n/a	n/a	100.0	4.010	n/a
S96T004421	D	Tungsten-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	< 4.010	<4.01e1	n/a	n/a	99.50	4.010	n/a
S96T004421	D	Uranium-ICP-Acid Dil.	ug/mL	97.20	<5.00e-01	< 80.20	<8.02e1	n/a	n/a	93.00	80.20	n/a
S96T004421	D	Vanadium-ICP-Acid Dil.	ug/mL	98.00	<2.00e-01	<2.00e+02	<2.00e2	n/a	n/a	98.00	200.0	n/a
S96T004421	D	Zinc-ICP-Acid Dil.	ug/mL	102.6	<1.00e-02	< 20.10	<2.01e1	62.90	9.22	98.00	20.10	n/a
S96T004421	D	Zirconium-ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	< 4.010	<4.01e1	n/a	n/a	103.0	4.010	n/a
S96T004421	D	Acetate by IC-Dioxex 4000/4500	ug/mL	100.8	<2.00e-02	4.30e+02	761.0	595.5	55.6	84.68	103.0	n/a
S96T004421	D	Formate by Ion Chromatograph	ug/mL	104.2	<2.10e-02	1.89e+03	1.80e+03	1.85e+03	4.88	112.5	108.2	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961004421		Fluoride-IC-Bionex 4000/4500	ug/mL	100.2	<1.30e-02	4.19e+02	407.0	412.9	2.91	n/a	34.14	n/a	
S961004421		Chloride-IC-Bionex 4000/4500	ug/mL	101.4	<1.70e-02	1.09e+04	1.10e+04	1.09e+04	0.91	n/a	44.54	n/a	
S961004421		Nitrite-IC-Bionex 4000/4500	ug/mL	95.56	<1.08e-01	1.20e+05	1.21e+05	1.20e+05	0.83	n/a	447.2	n/a	
S961004421		Bromide by Ion Chromatograph	ug/mL	98.64	<1.26e-01	1.03e+03	1.01e+03	1.02e+03	1.96	n/a	350.9	n/a	
S961004421		Nitrate by Ion Chromatograph	ug/mL	99.66	<1.59e-01	1.66e+05	1.68e+05	1.67e+05	1.20	n/a	373.6	n/a	
S961004421		Phosphate-IC-Bionex 4000/4500	ug/mL	99.45	<1.19e-01	6.81e+02	641.0	661.1	6.05	n/a	312.2	n/a	
S961004421		Sulfate by IC-Bionex 4000/4500	ug/mL	102.7	<1.36e-01	9.86e+02	952.0	968.8	3.51	n/a	356.9	n/a	
S961004421		Oxalate by IC-Bionex 4000/4500	ug/mL	100.4	<1.05e-01	<2.76e+02	313.0	n/a	n/a	n/a	273.7	n/a	
S961004421		Alpha in Liquid Samples	uCi/mL	80.47	<1.42e-02	<1.42e-02	<1.42e-2	n/a	n/a	77.44	3.60e-02	5.00E+02	

CORE NUMBER: 153
SEGMENT #: 8

Table 1. Data Summary Table
AN-105

SEGMENT POSITION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003975		DSC Exotherm using Mettler	Joules/g	96.31	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003975		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003975		Specific Gravity	Sp.G.	99.14	n/a	1.433	1.424	1.429	0.03	n/a	1.00e-02	n/a
S961003975		% Water by TGA using Mettler	%	99.48	n/a	49.99	50.16	50.08	0.34	n/a	1.00e-02	n/a
S961003975		0 Silver-TP-Acid Dtl.	ug/mL	100.0	<1.00e-02	17.30	15.80	16.55	9.06	84.70	4.01e	n/a
S961003975		0 Aluminum-TP-Acid Dtl.	ug/mL	98.00	<5.00e-02	3.57e+04	3.44e+04	3.50e+04	3.71	404.0	30.10	n/a
S961003975		0 Arsenic-TP-Acid Dtl.	ug/mL	102.8	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	88.90	40.10	n/a
S961003975		0 Boron-TP-Acid Dtl.	ug/mL	99.40	<5.00e-02	<67.40	<6.74e0	<6.74e0	1.80	92.80	30.10	n/a
S961003975		0 Barium-TP-Acid Dtl.	ug/mL	100.0	<5.00e-02	<20.10	<2.01e1	<2.01e1	n/a	92.80	30.10	n/a
S961003975		0 Beryllium-TP-Acid Dtl.	ug/mL	102.2	<5.00e-03	<2.00e	<2.00e0	<2.00e0	n/a	92.80	30.10	n/a
S961003975		0 Bismuth-TP-Acid Dtl.	ug/mL	101.8	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Calcium-TP-Acid Dtl.	ug/mL	100.4	<5.00e-03	<2.00e	<2.00e0	<2.00e0	n/a	92.80	30.10	n/a
S961003975		0 Cadmium-TP-Acid Dtl.	ug/mL	100.4	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Cobalt-TP-Acid Dtl.	ug/mL	100.4	<1.00e-03	<2.00e	<2.00e0	<2.00e0	n/a	92.80	30.10	n/a
S961003975		0 Chromium-TP-Acid Dtl.	ug/mL	101.4	<2.00e-02	<8.02e0	<8.02e0	<8.02e0	n/a	92.80	30.10	n/a
S961003975		0 Copper-TP-Acid Dtl.	ug/mL	103.4	<1.00e-02	<4.01e	<4.01e0	<4.01e0	n/a	92.80	30.10	n/a
S961003975		0 Iron-TP-Acid Dtl.	ug/mL	101.0	<5.00e-02	<20.10	<2.01e1	<2.01e1	n/a	92.80	30.10	n/a
S961003975		0 Potassium-TP-Acid Dtl.	ug/mL	96.60	<5.00e-01	<6.33e+03	5.90e+03	6.12e+03	7.03	92.80	30.10	n/a
S961003975		0 Lanthanum-TP-Acid Dtl.	ug/mL	102.2	<5.00e-02	<20.10	<2.01e1	<2.01e1	n/a	92.80	30.10	n/a
S961003975		0 Lithium-TP-Acid Dtl.	ug/mL	101.6	<1.00e-02	<7.90e0	<7.90e0	<7.90e0	n/a	92.80	30.10	n/a
S961003975		0 Magnesium-TP-Acid Dtl.	ug/mL	97.40	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Manganese-TP-Acid Dtl.	ug/mL	98.40	<1.00e-02	<4.01e	<4.01e0	<4.01e0	n/a	92.80	30.10	n/a
S961003975		0 Molybdenum-TP-Acid Dtl.	ug/mL	100.8	<5.00e-02	<91.00	<9.10e0	<9.10e0	n/a	92.80	30.10	n/a
S961003975		0 Sodium-TP-Acid Dtl.	ug/mL	98.80	<1.00e-01	<2.39e+05	2.23e+05	2.31e+05	6.93	92.80	30.10	n/a
S961003975		0 Neodymium-TP-Acid Dtl.	ug/mL	102.6	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Nickel-TP-Acid Dtl.	ug/mL	100.2	<2.00e-02	<3.54e+02	3.51e0	3.52e5	0.85	92.80	30.10	n/a
S961003975		0 Phosphorus-TP-Acid Dtl.	ug/mL	101.8	<2.00e-01	<48.60	<4.86e0	<4.86e0	10.6	92.80	30.10	n/a
S961003975		0 Lead-TP-Acid Dtl.	ug/mL	100.6	<1.00e-01	<6.57e+02	6.17e0	6.37e0	6.28	92.80	30.10	n/a
S961003975		0 Sulfur-TP-Acid Dtl.	ug/mL	99.40	<6.00e-02	<24.10	<2.41e1	<2.41e1	n/a	92.80	30.10	n/a
S961003975		0 Antimony-TP-Acid Dtl.	ug/mL	97.20	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Selenium-TP-Acid Dtl.	ug/mL	99.60	<5.00e-02	<2.07e+02	2.02e0	2.04e5	2.44	92.80	30.10	n/a
S961003975		0 Silicon-TP-Acid Dtl.	ug/mL	98.80	<1.00e-01	<40.10	<4.01e1	<4.01e1	n/a	92.80	30.10	n/a
S961003975		0 Samarium-TP-Acid Dtl.	ug/mL	99.60	<1.00e-02	<4.01e	<4.01e0	<4.01e0	n/a	92.80	30.10	n/a
S961003975		0 Strontium-TP-Acid Dtl.	ug/mL	98.00	<1.00e-02	<4.01e	<4.01e0	<4.01e0	n/a	92.80	30.10	n/a
S961003975		0 Thallium-TP-Acid Dtl.	ug/mL	98.00	<2.00e-01	<80.20	<8.02e0	<8.02e0	n/a	92.80	30.10	n/a
S961003975		0 Uranium-TP-Acid Dtl.	ug/mL	96.40	<5.00e-01	<2.00e+02	<2.00e2	<2.00e2	n/a	92.80	30.10	n/a
S961003975		0 Vanadium-TP-Acid Dtl.	ug/mL	100.6	<1.00e-02	<20.10	<2.01e1	<2.01e1	n/a	92.80	30.10	n/a
S961003975		0 Zinc-TP-Acid Dtl.	ug/mL	101.4	<1.00e-02	<57.50	<5.75e0	<5.75e0	33.3	92.80	30.10	n/a
S961003975		0 Zirconium-TP-Acid Dtl.	ug/mL	99.20	<1.00e-02	<4.34e0	<4.34e0	<4.34e0	n/a	92.80	30.10	n/a
S961003975		Acetate by IC-Dionex 4000/4500	ug/mL	100.8	<2.00e-02	<2.37e+02	2.45e0	2.41e3	n/a	92.80	30.10	n/a
S961003975		Formate by Ion Chromatograph	ug/mL	91.67	<2.10e-02	5.95e+02	5.99e0	5.66e1	9.54	n/a	54.62	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003975		Fluoride-IC-Dionex 4000/4500	ug/mL	106.3	<1.30e-02	4.40e+02	413.0	426.6	6.33	88.64	66.96	n/a
S96T003975		Chloride-IC-Dionex 4000/4500	ug/mL	106.8	3.30e-02	9.58e+03	9.53e+03	9.56e+03	0.52	111.0	87.57	n/a
S96T003975		Nitrite-IC - Dionex 4000/4500	ug/mL	100.8	<1.07e-01	1.21e+03	1.21e+03	1.21e+03	0.00	105.6	551.2	n/a
S96T003975		Bromide by Ion Chromatograph	ug/mL	98.81	<1.26e-01	1.17e+03	1.17e+03	1.17e+03	0.00	96.94	649.0	n/a
S96T003975		Nitrate by IC-Dionex 4000/4500	ug/mL	102.7	1.74e-01	1.60e+03	1.59e+03	1.59e+03	0.63	118.7	721.1	n/a
S96T003975		Phosphate-IC-Dionex 4000/4500	ug/mL	99.82	<1.19e-01	<6.12e+02	<6.12e+02	n/a	n/a	105.1	612.5	n/a
S96T003975		Sulfate by IC-Dionex 4000/4500	ug/mL	102.2	<1.36e-01	1.14e+03	1.27e+03	1.20e+03	10.8	103.5	700.0	n/a
S96T003975		Oxalate by IC-Dionex 4000/4500	ug/mL	94.12	<1.05e-01	<5.41e+02	582.0	n/a	n/a	97.53	540.9	n/a
S96T003975		Alpha in Liquid Samples	uCi/mL	93.75	<2.94e-02	<2.94e-02	<1.45E-1	n/a	n/a	90.25	3.90e-02	2.45E+02

Table 1. Data Summary Table
AM-105

CORE NUMBER: 153
SEGMENT #: 9

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003976		DSC Exotherm using Mettler	Joules/g	106.9	n/a	2.200	0.00e+00	1.100	n/a	n/a	n/a	n/a
S961003976		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	4.440	0.00e+00	2.220	200	n/a	n/a	n/a
S961003976		% Water by TGA using Mettler	%	99.14	n/a	1.476	1.476	1.459	2.33	n/a	1.00e-02	n/a
S961003976		D Silver-ICP-Acid DIL.	ug/ml	100.0	<1.00e-02	50.51	50.33	50.42	2.33	n/a	1.00e-02	n/a
S961003976		D Aluminum-ICP-Acid DIL.	ug/ml	98.00	<3.00e-02	16.10	16.60	17.35	14.4	86.70	4.010	n/a
S961003976		D Arsenic-ICP-Acid DIL.	ug/ml	102.8	<1.00e-01	3.73e+04	4.44e+04	4.08e+04	17.4	1.01e+03	20.10	n/a
S961003976		D Boron-ICP-Acid DIL.	ug/ml	102.8	<3.00e-02	66.10	66.10	73.40	19.9	95.40	40.10	n/a
S961003976		D Barium-ICP-Acid DIL.	ug/ml	102.0	<3.00e-02	2.00	2.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Beryllium-ICP-Acid DIL.	ug/ml	102.2	<3.00e-03	20.10	2.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Bismuth-ICP-Acid DIL.	ug/ml	101.8	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Calcium-ICP-Acid DIL.	ug/ml	100.4	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Cadmium-ICP-Acid DIL.	ug/ml	100.8	<3.00e-03	2.000	2.00e0	73.40	19.9	95.40	20.10	n/a
S961003976		D Cerium-ICP-Acid DIL.	ug/ml	100.4	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Cobalt-ICP-Acid DIL.	ug/ml	101.4	<2.00e-02	8.020	8.02e0	73.40	19.9	95.40	20.10	n/a
S961003976		D Chromium-ICP-Acid DIL.	ug/ml	101.4	<1.00e-02	1.90e+02	2.25e+02	207.5	16.9	103.0	4.010	n/a
S961003976		D Copper-ICP-Acid DIL.	ug/ml	103.4	<1.00e-02	4.010	4.01e0	73.40	19.9	95.40	20.10	n/a
S961003976		D Iron-ICP-Acid DIL.	ug/ml	101.0	<5.00e-02	20.10	2.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Potassium-ICP-Acid DIL.	ug/ml	96.60	<5.00e-01	5.81e+03	6.36e+03	6.36e+03	17.2	233.0	200.0	n/a
S961003976		D Lanthanum-ICP-Acid DIL.	ug/ml	102.2	<5.00e-02	20.10	2.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Magnesium-ICP-Acid DIL.	ug/ml	101.6	<1.00e-02	4.010	4.15e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Manganese-ICP-Acid DIL.	ug/ml	97.40	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Molybdenum-ICP-Acid DIL.	ug/ml	98.40	<1.00e-02	4.010	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Sodium-ICP-Acid DIL.	ug/ml	100.8	<5.00e-02	83.00	99.50	91.25	18.1	101.0	20.10	n/a
S961003976		D Neodymium-ICP-Acid DIL.	ug/ml	98.80	<1.00e-01	2.18e+03	2.58e+03	2.38e+03	16.8	5.30e+03	40.10	n/a
S961003976		D Nickel-ICP-Acid DIL.	ug/ml	102.6	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Phosphorus-ICP-Acid DIL.	ug/ml	100.2	<2.00e-02	8.020	8.02e0	73.40	19.9	95.40	20.10	n/a
S961003976		D Lead-ICP-Acid DIL.	ug/ml	101.8	<2.00e-01	3.75e+02	4.35e+02	405.0	14.8	146.0	80.20	n/a
S961003976		D Sulfur-ICP-Acid DIL.	ug/ml	100.6	<1.00e-01	48.90	50.00	49.45	2.22	99.10	40.10	n/a
S961003976		D Antimony-ICP-Acid DIL.	ug/ml	99.40	<1.00e-01	8.74e+02	1.04e+03	957.0	17.3	125.0	40.10	n/a
S961003976		D Selenium-ICP-Acid DIL.	ug/ml	96.20	<6.00e-01	24.10	2.41e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Silicon-ICP-Acid DIL.	ug/ml	97.20	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Strontium-ICP-Acid DIL.	ug/ml	99.60	<5.00e-02	2.01e+02	2.44e+02	222.5	19.3	106.0	20.10	n/a
S961003976		D Titanium-ICP-Acid DIL.	ug/ml	98.80	<1.00e-01	40.10	4.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Vanadium-ICP-Acid DIL.	ug/ml	98.60	<1.00e-02	4.010	4.01e0	73.40	19.9	95.40	20.10	n/a
S961003976		D Uranium-ICP-Acid DIL.	ug/ml	98.00	<2.00e-01	80.20	8.02e1	73.40	19.9	95.40	20.10	n/a
S961003976		D Zirconium-ICP-Acid DIL.	ug/ml	96.40	<5.00e-01	2.00e+02	2.40e+02	222.5	19.3	106.0	20.10	n/a
S961003976		D Zirconium-ICP-Acid DIL.	ug/ml	100.6	<5.00e-02	50.10	5.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		Accelerate by IC-Dioxin 4000/4500	ug/ml	101.4	<1.00e-02	50.10	5.01e1	73.40	19.9	95.40	20.10	n/a
S961003976		Formate by Ion Chromatograph	ug/ml	99.20	<2.00e-02	4.010	4.01e0	73.40	19.9	95.40	20.10	n/a
S961003976			ug/ml	111.3	<2.00e-02	6.74e+03	6.96e+03	684.9	3.21	n/a	103.0	n/a
S961003976			ug/ml	98.33	<2.10e-02	1.73e+03	1.74e+03	1.74e+03	1.72	n/a	108.2	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S941003976		Fluoride-IC-Dionex 4000/4500	ug/ml	106.3	<1.30e-02	4.60e+02	429.0	444.4	6.97	n/a	27.57	n/a
S941003976		Chloride-IC-Dionex 4000/4500	ug/ml	106.8	3.30e-02	9.45e+03	9.62e+03	9.53e+03	1.78	n/a	36.06	n/a
S941003976		Nitrite-IC-Dionex 4000/4500	ug/ml	100.8	<1.07e-01	1.17e+05	1.18e+05	1.18e+05	0.85	n/a	551.2	n/a
S941003976		Bromide by Ion Chromatograph	ug/ml	98.81	<1.26e-01	6.52e+02	673.0	662.5	3.17	n/a	267.2	n/a
S941003976		Nitrate by IC-Dionex 4000/4500	ug/ml	102.7	<1.74e-01	1.55e+05	1.53e+05	1.54e+05	1.30	n/a	721.1	n/a
S941003976		Phosphate-IC-Dionex 4000/4500	ug/ml	99.82	<1.19e-01	7.51e+02	783.0	767.1	4.17	n/a	252.2	n/a
S941003976		Sulfate by IC-Dionex 4000/4500	ug/ml	102.2	<1.36e-01	8.42e+02	830.0	835.9	1.44	n/a	288.2	n/a
S941003976		Oxalate by IC-Dionex 4000/4500	ug/ml	94.12	<1.05e-01	2.70e+02	274.0	271.9	1.47	n/a	222.7	n/a
S961003976		Alpha in Liquid Samples	ug/ml	93.75	<2.94e-02	4.48e-02	<2.94E-2	n/a	n/a	n/a	3.90e-02	8.26E+01

A-0002-3

CORE NUMBER: 153

SEGMENT #: 10

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004309		DSC Exotherm Dry	Joules/g Dry	n/a	n/a	26.50	38.20	32.35	36.2	n/a	n/a	n/a
S96T004309		DSC Exotherm on Perkin Elmer	Joules/g	96.59	n/a	13.55	19.54	16.55	36.2	n/a	n/a	n/a
S96T004309		Specific Gravity	Sp. G.	97.00	n/a	1.364	1.464	1.414	7.07	n/a	1.00e-02	n/a
S96T004309		% Water by IGA on Perkin Elmer	%	99.09	n/a	48.12	49.52	48.82	2.87	n/a	n/a	n/a
S96T004309		Aluminum-ICP-Acid Dil.	ug/mL	100.0	<1.00e-02	17.10	18.80	17.95	9.47	92.00	6.010	n/a
S96T004309		Silver-ICP-Acid Dil.	ug/mL	98.00	<5.00e-02	4.32e+04	4.50e+04	4.41e+04	4.08	24.90	30.10	n/a
S96T004309		Arsenic-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	96.40	60.10	n/a
S96T004309		Boron-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	73.20	77.70	75.45	5.96	98.70	30.10	n/a
S96T004309		Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	96.50	30.10	n/a
S96T004309		Beryllium-ICP-Acid Dil.	ug/mL	104.0	<5.00e-03	< 3.00	<3.00e0	n/a	n/a	102.0	<3.000	n/a
S96T004309		Bismuth-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	103.0	60.10	n/a
S96T004309		Calcium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a
S96T004309		Cadmium-ICP-Acid Dil.	ug/mL	101.2	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	99.20	<3.000	n/a
S96T004309		Cerium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	97.00	60.10	n/a
S96T004309		Cobalt-ICP-Acid Dil.	ug/mL	102.2	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	99.60	12.00	n/a
S96T004309		Chromium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-02	2.17e+02	229.0	223.0	5.38	100.0	6.010	n/a
S96T004309		Copper-ICP-Acid Dil.	ug/mL	97.00	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	94.20	6.010	n/a
S96T004309		Iron-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	99.60	30.10	n/a
S96T004309		Potassium-ICP-Acid Dil.	ug/mL	98.80	<5.00e-01	6.64e+03	6.93e+03	6.78e+03	4.27	74.80	300.0	n/a
S96T004309		Lithium-ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	98.50	30.10	n/a
S96T004309		Magnesium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	102.0	6.010	n/a
S96T004309		Manganese-ICP-Acid Dil.	ug/mL	97.60	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	95.50	60.10	n/a
S96T004309		Molybdenum-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	94.80	6.010	n/a
S96T004309		Sodium-ICP-Acid Dil.	ug/mL	102.0	<5.00e-02	1.01e+02	105.0	103.0	3.88	99.70	30.10	n/a
S96T004309		Niobium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	2.54e+05	2.66e+05	2.60e+05	4.62	-4.24e+02	60.10	n/a
S96T004309		Nickel-ICP-Acid Dil.	ug/mL	100.4	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	98.70	60.10	n/a
S96T004309		Phosphorus-ICP-Acid Dil.	ug/mL	101.2	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	99.10	12.00	n/a
S96T004309		Lead-ICP-Acid Dil.	ug/mL	100.8	<1.00e-01	3.66e+02	465.0	415.5	23.8	106.0	120.0	n/a
S96T004309		Sulfur-ICP-Acid Dil.	ug/mL	100.8	<1.00e-01	1.02e+03	1.05e+03	1.04e+03	2.90	102.0	60.10	n/a
S96T004309		Antimony-ICP-Acid Dil.	ug/mL	97.20	<5.00e-02	< 36.10	<3.61e1	n/a	n/a	96.20	36.10	n/a
S96T004309		Selenium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	106.0	60.10	n/a
S96T004309		Strontium-ICP-Acid Dil.	ug/mL	94.80	<5.00e-02	1.95e+02	209.0	202.0	6.93	96.90	30.10	n/a
S96T004309		Titanium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	96.10	60.10	n/a
S96T004309		Tantalum-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	96.50	6.010	n/a
S96T004309		Thallium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	92.10	120.0	n/a
S96T004309		Vanadium-ICP-Acid Dil.	ug/mL	95.40	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	96.00	6.010	n/a
S96T004309		Zinc-ICP-Acid Dil.	ug/mL	95.90	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	96.00	6.010	n/a
S96T004309		Zirconium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	99.40	30.10	n/a
S96T004309		Zinc-ICP-Acid Dil.	ug/mL	102.4	<1.00e-02	< 60.10	<6.01e1	n/a	n/a	96.00	300.0	n/a
S96T004309		Zinc-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	97.80	6.010	n/a
S96T004309		Acetate by IC-Dioxex 4000/4500	ug/mL	104.0	<2.00e-02	8.12e+02	858.0	835.2	5.51	79.60	103.0	n/a
S96T004309		Formate by Ion Chromatograph	ug/mL	105.8	<2.10e-02	2.04e+03	1.98e+03	2.01e+03	2.99	108.3	108.2	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004309		Fluoride-IC-Dionex 4000/4500	ug/mL	99.32	<1.30e-02	4.33e+02	408.0	420.5	5.95	n/a	53.83	n/a
S96T004309		Chloride-IC-Dionex 4000/4500	ug/mL	103.0	<1.70e-02	9.32e+03	8.96e+03	9.14e+03	3.94	n/a	70.40	n/a
S96T004309		Nitrite-IC - Dionex 4000/4500	ug/mL	101.1	<1.07e-01	1.13e+05	1.11e+05	1.13e+05	3.54	n/a	443.1	n/a
S96T004309		Bromide by Ion Chromatograph	ug/mL	99.49	<1.26e-01	9.93e+02	955.0	974.1	3.90	n/a	521.8	n/a
S96T004309		Nitrate by IC-Dionex 4000/4500	ug/mL	104.4	1.50e-01	1.54e+05	1.49e+05	1.52e+05	3.30	n/a	579.7	n/a
S96T004309		Phosphate-IC-Dionex 4000/4500	ug/mL	101.3	<1.19e-01	6.57e+02	565.0	611.0	15.1	n/a	492.4	n/a
S96T004309		Sulfate by IC-Dionex 4000/4500	ug/mL	103.3	<1.36e-01	1.09e+03	1.04e+03	1.06e+03	4.69	n/a	562.8	n/a
S96T004309		Oxalate by IC-Dionex 4000/4500	ug/mL	101.3	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	n/a	434.8	n/a
S96T004309		Alpha in Liquid Samples	uCi/mL	95.31	1.10e-02	<7.10e-03	<7.10E-3	n/a	n/a	n/a	1.80e-02	5.00E+02

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S961003979	D	DSC Exotherm DRY	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S961003979	D	DSC Exotherm on Perkin Elmer	Joules/g	104.-4	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S961003979	D	Specific Gravity	Sp.G.	99.34	n/a	1.464	1.445	1.454	1.31	1.00e-02	n/a
S961003979	D	% Water by TGA on Perkin Elmer	%	99.34	n/a	50.03	50.06	50.05	0.06	n/a	n/a
S961003979	D	Silver-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	19.00	18.20	18.60	4.30	90.60	n/a
S961003979	D	Aluminum-ICP-Acid Dil.	ug/mL	98.40	<5.00e-02	4.29e+04	4.27e+04	4.28e+04	0.47	11.00	n/a
S961003979	D	Arsenic-ICP-Acid Dil.	ug/mL	104.-6	<1.00e-01	< 60.10	< 60.10	n/a	n/a	30.10	n/a
S961003979	D	Boron-ICP-Acid Dil.	ug/mL	100.-0	<5.00e-02	74.50	77.70	76.10	4.20	96.60	n/a
S961003979	D	Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 30.10	< 30.10	n/a	n/a	30.10	n/a
S961003979	D	Beryllium-ICP-Acid Dil.	ug/mL	102.-6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	3.000	n/a
S961003979	D	Bismuth-ICP-Acid Dil.	ug/mL	101.-8	<1.00e-01	< 60.10	< 60.10	n/a	n/a	60.10	n/a
S961003979	D	Calcium-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	< 60.10	< 60.10	n/a	n/a	3.000	n/a
S961003979	D	Cadmium-ICP-Acid Dil.	ug/mL	100.-6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	99.20	n/a
S961003979	D	Cerium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	< 60.10	< 60.10	n/a	n/a	99.50	n/a
S961003979	D	Cobalt-ICP-Acid Dil.	ug/mL	101.-0	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	98.50	n/a
S961003979	D	Chromium-ICP-Acid Dil.	ug/mL	100.-8	<1.00e-02	2.15e+02	211.0	213.0	1.88	6.010	n/a
S961003979	D	Copper-ICP-Acid Dil.	ug/mL	103.-6	<1.00e-02	< 6.010	< 6.01e0	n/a	n/a	6.010	n/a
S961003979	D	Iron-ICP-Acid Dil.	ug/mL	101.-0	<5.00e-02	< 30.10	< 30.10	n/a	n/a	30.10	n/a
S961003979	D	Potassium-ICP-Acid Dil.	ug/mL	100.-2	<5.00e-01	6.92e+03	6.87e+03	6.80e+03	3.68	64.00	n/a
S961003979	D	Lanthanum-ICP-Acid Dil.	ug/mL	101.-4	<5.00e-02	< 30.10	< 30.10	n/a	n/a	100.0	n/a
S961003979	D	Lithium-ICP-Acid Dil.	ug/mL	102.-0	<1.00e-02	< 60.10	< 60.10	n/a	n/a	103.0	n/a
S961003979	D	Magnesium-ICP-Acid Dil.	ug/mL	96.60	<1.00e-01	< 60.10	< 60.10	n/a	n/a	92.90	n/a
S961003979	D	Manganese-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	< 6.010	< 6.01e0	n/a	n/a	6.010	n/a
S961003979	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.-8	<5.00e-02	< 60.60	95.80	96.20	0.83	99.50	n/a
S961003979	D	Sodium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	2.52e+05	2.50e+05	2.51e+05	0.80	-6.150e2	n/a
S961003979	D	Niobium-ICP-Acid Dil.	ug/mL	101.-4	<1.00e-01	< 60.10	< 60.10	n/a	n/a	100.0	n/a
S961003979	D	Nickel-ICP-Acid Dil.	ug/mL	99.80	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	99.40	n/a
S961003979	D	Phosphorus-ICP-Acid Dil.	ug/mL	100.-4	<2.00e-01	3.49e+02	343.0	346.0	1.73	99.50	n/a
S961003979	D	Lead-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	< 60.10	< 60.10	n/a	n/a	105.0	n/a
S961003979	D	Sulfur-ICP-Acid Dil.	ug/mL	95.80	<6.00e-02	< 36.10	< 3.61e1	980.5	1.94	95.00	n/a
S961003979	D	Antimony-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	< 60.10	< 60.10	n/a	n/a	94.70	n/a
S961003979	D	Selenium-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	< 60.10	< 60.10	n/a	n/a	98.20	n/a
S961003979	D	Silicon-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	2.57e+02	233.0	235.0	1.70	98.50	n/a
S961003979	D	Samarium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	< 60.10	< 60.10	n/a	n/a	96.40	n/a
S961003979	D	Strontium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	< 6.010	< 6.01e0	n/a	n/a	97.00	n/a
S961003979	D	Thallium-ICP-Acid Dil.	ug/mL	96.20	<1.00e-01	< 60.10	< 60.10	n/a	n/a	95.80	n/a
S961003979	D	Thallium-ICP-Acid Dil.	ug/mL	96.10	<5.00e-01	<3.20e+02	<3.00e2	n/a	n/a	120.0	n/a
S961003979	D	Uranium-ICP-Acid Dil.	ug/mL	100.-2	<5.00e-02	< 30.10	< 30.10	n/a	n/a	95.00	n/a
S961003979	D	Vanadium-ICP-Acid Dil.	ug/mL	101.-2	<1.00e-02	< 14.50	13.90	14.20	4.23	99.50	n/a
S961003979	D	Zinc-ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	< 6.010	< 6.01e0	n/a	n/a	98.10	n/a
S961003979	D	Zirconium-ICP-Acid Dil.	ug/mL	109.-7	<2.00e-02	6.69e+02	655.0	662.0	2.11	86.29	n/a
S961003979	D	Acetate by IC-Dioxen 4000/4500	ug/mL	100.-0	<2.10e-02	1.70e+03	1.77e+03	1.73e+03	4.03	105.-8	n/a
S961003979	D	Formate by Ion Chromatograph	ug/mL	100.-0	<2.10e-02	1.70e+03	1.77e+03	1.73e+03	4.03	105.-8	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003979			Fluoride-IC-Dionex 4000/4500	ug/mL	103.2	<1.30e-02	< 55.83	<5.38e1	n/a	n/a	153.4	55.83	n/a
S96T003979			Chloride-IC-Dionex 4000/4500	ug/mL	99.49	<1.70e-01	9.64e+03	9.58e+03	9.61e+03	0.62	101.6	70.40	n/a
S96T003979			Nitrite-IC - Dionex 4000/4500	ug/mL	98.35	<1.07e-01	1.22e+05	1.24e+05	1.23e+05	1.63	120.4	443.1	n/a
S96T003979			Bromide by Ion Chromatograph	ug/mL	98.30	<1.26e-01	6.70e+02	678.0	674.0	1.19	94.91	521.8	n/a
S96T003979			Nitrate by IC-Dionex 4000/4500	ug/mL	99.30	1.80e-01	1.53e+05	1.53e+05	1.53e+05	0.00	120.0	579.7	n/a
S96T003979			Phosphate-IC-Dionex 4000/4500	ug/mL	97.43	<1.19e-01	8.12e+02	619.0	715.6	27.0	99.27	492.4	n/a
S96T003979			Sulfate by IC-Dionex 4000/4500	ug/mL	99.53	<1.36e-01	1.05e+03	1.03e+03	1.04e+03	1.92	98.89	562.8	n/a
S96T003979			Oxalate by IC-Dionex 4000/4500	ug/mL	100.6	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	102.3	434.8	n/a
S96T003979			Alpha In Liquid Samples	uCi/mL	93.75	<2.94e-02	<2.58e-02	<3.29e-2	n/a	n/a	n/a	3.90e-02	5.00e+02

Table 1. Data Summary Table
AN-105

CORE NUMBER: 153

SEGMENT #: 12

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003977		DSC Exotherm using Mettler	Joules/g	106.9	n/a	0.00e+00	52.40	26.20	200	n/a	n/a	n/a
S96T003977		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	105.0	52.50	200	n/a	n/a	n/a
S96T003977		Specific Gravity	Sp.G.	99.14	n/a	1.462	1.427	1.445	2.42	n/a	1.00e-02	n/a
S96T003977		% Water by TGA using Mettler	%	99.53	n/a	49.68	50.25	49.97	1.14	n/a	n/a	n/a
S96T003977	D	Silver-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	17.40	18.60	18.00	6.67	90.20	6.010	n/a
S96T003977	D	Aluminum-ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	4.24e+04	4.34e+04	4.28e+04	2.57	308.0	30.10	n/a
S96T003977	D	Arsenic-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	<60.10	<6.01e1	<60.10	n/a	95.60	60.10	n/a
S96T003977	D	Boron-ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	76.20	76.20	75.90	0.79	97.10	30.10	n/a
S96T003977	D	Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	<30.10	<3.01e1	n/a	n/a	96.80	30.10	n/a
S96T003977	D	Beryllium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-03	<3.000	<3.00e0	n/a	n/a	99.90	3.000	n/a
S96T003977	D	Bismuth-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	<60.10	<6.01e1	n/a	n/a	95.80	60.10	n/a
S96T003977	D	Calcium-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	<60.10	<6.01e1	n/a	n/a	99.70	60.10	n/a
S96T003977	D	Cadmium-ICP-Acid Dil.	ug/mL	100.6	<5.00e-03	<3.000	<3.00e0	n/a	n/a	98.80	3.000	n/a
S96T003977	D	Cerium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	<60.10	<6.01e1	n/a	n/a	98.80	60.10	n/a
S96T003977	D	Cobalt-ICP-Acid Dil.	ug/mL	101.0	<2.00e-02	<12.00	<1.20e1	n/a	n/a	98.80	12.00	n/a
S96T003977	D	Chromium-ICP-Acid Dil.	ug/mL	100.8	<1.00e-02	2.10e+02	218.0	214.0	3.74	100.0	6.010	n/a
S96T003977	D	Copper-ICP-Acid Dil.	ug/mL	103.6	<1.00e-02	<6.010	<6.01e0	n/a	n/a	100.0	6.010	n/a
S96T003977	D	Iron-ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	<30.10	<3.01e1	n/a	n/a	98.60	30.10	n/a
S96T003977	D	Potassium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-01	6.68e+03	6.92e+03	6.80e+03	3.53	152.0	300.0	n/a
S96T003977	D	Lanthanum-ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	<30.10	<3.01e1	n/a	n/a	100.0	30.10	n/a
S96T003977	D	Lithium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	<6.010	<6.01e0	n/a	n/a	103.0	6.010	n/a
S96T003977	D	Magnesium-ICP-Acid Dil.	ug/mL	96.60	<1.00e-01	<60.10	<6.01e1	n/a	n/a	91.80	60.10	n/a
S96T003977	D	Manganese-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	<6.010	<6.01e0	n/a	n/a	94.20	6.010	n/a
S96T003977	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	96.00	101.0	98.50	5.08	99.90	30.10	n/a
S96T003977	D	Sodium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	2.52e+05	2.58e+05	2.55e+05	2.35	1.14e+03	60.10	n/a
S96T003977	D	Neodymium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	<60.10	<6.01e1	n/a	n/a	99.90	60.10	n/a
S96T003977	D	Nickel-ICP-Acid Dil.	ug/mL	99.80	<2.00e-02	<12.00	<1.20e1	n/a	n/a	97.60	12.00	n/a
S96T003977	D	Phosphorus-ICP-Acid Dil.	ug/mL	100.4	<2.00e-01	3.60e+02	359.0	359.5	0.28	96.70	120.0	n/a
S96T003977	D	Lead-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	<60.10	<6.01e1	n/a	n/a	106.0	60.10	n/a
S96T003977	D	Sulfur-ICP-Acid Dil.	ug/mL	98.80	<1.00e-02	9.78e+02	1.00e+03	989.0	2.22	103.0	60.10	n/a
S96T003977	D	Antimony-ICP-Acid Dil.	ug/mL	95.80	<5.00e-02	<36.10	<3.61e1	n/a	n/a	99.60	36.10	n/a
S96T003977	D	Selenium-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	<60.10	<6.01e1	n/a	n/a	99.90	60.10	n/a
S96T003977	D	Samarium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	2.31e+02	248.0	239.5	7.10	103.0	30.10	n/a
S96T003977	D	Strontium-ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	<6.010	<6.01e1	n/a	n/a	95.90	6.010	n/a
S96T003977	D	Titanium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	<6.010	<6.01e0	n/a	n/a	96.70	6.010	n/a
S96T003977	D	Thallium-ICP-Acid Dil.	ug/mL	96.20	<1.00e-02	<1.20e+02	<1.20e2	n/a	n/a	92.30	120.0	n/a
S96T003977	D	Uranium-ICP-Acid Dil.	ug/mL	96.10	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	95.50	300.0	n/a
S96T003977	D	Vanadium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	<30.10	<3.01e1	n/a	n/a	98.30	30.10	n/a
S96T003977	D	Zinc-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	15.80	16.30	16.05	3.12	100.0	6.010	n/a
S96T003977	D	Zirconium-ICP-Acid Dil.	ug/mL	98.60	<2.00e-02	<6.010	<6.01e0	n/a	n/a	98.00	6.010	n/a
S96T003977	D	Acetate by IC-Dioxon 4000/4500	ug/mL	111.3	<2.00e-02	6.18e+02	691.0	654.7	11.2	n/a	103.0	n/a
S96T003977	D	Formate by Ion Chromatograph	ug/mL	98.33	<2.10e-02	1.86e+03	1.77e+03	1.81e+03	4.96	n/a	108.2	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003977		Fluoride-IC-Dionex 4000/4500	ug/mL	100.7	<1.30e-02	< 53.83	<5.38e1	n/a	n/a	154.4	53.83	n/a
S96T003977		Chloride-IC-Dionex 4000/4500	ug/mL	96.08	<1.70e-02	9.87e+03	1.02e+04	1.00e+04	3.29	107.0	70.40	n/a
S96T003977		Nitrite-IC - Dionex 4000/4500	ug/mL	95.59	<1.07e-01	1.26e+03	1.31e+05	1.28e+05	3.89	111.6	443.1	n/a
S96T003977		Bromide by Ion Chromatograph	ug/mL	99.49	<1.26e-01	7.93e+02	836.0	814.6	5.28	95.42	521.8	n/a
S96T003977		Nitrate by IC-Dionex 4000/4500	ug/mL	97.14	1.50e-01	1.68e+05	1.72e+05	1.70e+05	2.35	92.09	579.7	n/a
S96T003977		Phosphate-IC-Dionex 4000/4500	ug/mL	95.23	<1.19e-01	5.29e+02	773.0	650.8	37.5	97.61	492.4	n/a
S96T003977		Sulfate by IC-Dionex 4000/4500	ug/mL	97.46	<1.36e-01	1.00e+03	1.13e+03	1.07e+03	12.2	99.05	562.8	n/a
S96T003977		Oxalate by IC-Dionex 4000/4500	ug/mL	99.43	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	102.6	434.8	n/a
S96T003977		Alpha In Liquid Samples	uCi/mL	93.75	<2.94e-02	<2.30e-02	<3.29E-2	n/a	n/a	n/a	3.90e-02	3.59E+02

Table 1. Data Summary Table
AN-105

CORE NUMBER: 153
SEGMENT #: 13

SEGMENT PORTION: Drainable Liquid

Sample#	RI#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err#
S96T003978		DSC Exotherm Dry Calculated	Joules/g Dry	104.4	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003978		DSC Exotherm on Perkin Elmer	Joules/g	104.4	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003978		Specific Gravity	Sp. G.	99.34	n/a	1.472	1.478	1.475	0.41	n/a	1.00e-02	n/a
S96T003978		% Water by TGA on Perkin Elmer	%	99.34	n/a	49.41	50.44	49.92	2.06	n/a	n/a	n/a
S96T003978		D Silver-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	16.90	15.80	16.35	6.73	90.30	6.010	n/a
S96T003978		D Aluminum-ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	4.01e+04	3.70e+04	3.86e+04	8.04	111.0	30.10	n/a
S96T003978		D Arsenic-ICP-Acid Dil.	ug/mL	104.6	<1.00e-01	< 60.10	<60.10	< 60.10	n/a	95.90	60.10	n/a
S96T003978		D Boron-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	71.20	64.40	67.80	10.0	96.70	30.10	n/a
S96T003978		D Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	96.90	30.10	n/a
S96T003978		D Beryllium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	99.90	3.000	n/a
S96T003978		D Bismuth-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	< 60.10	<60.10	n/a	n/a	97.30	60.10	n/a
S96T003978		D Cadmium-ICP-Acid Dil.	ug/mL	99.20	<5.00e-03	< 3.000	<3.00e0	n/a	n/a	99.10	3.000	n/a
S96T003978		D Cerium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	< 60.10	<60.10	n/a	n/a	99.90	60.10	n/a
S96T003978		D Cobalt-ICP-Acid Dil.	ug/mL	101.0	<2.00e-02	< 12.00	<1.20e1	191.0	8.38	99.30	6.010	n/a
S96T003978		D Chromium-ICP-Acid Dil.	ug/mL	103.6	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	100.0	6.010	n/a
S96T003978		D Copper-ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	< 30.10	<3.01e1	6.06e+03	7.42	98.80	30.10	n/a
S96T003978		D Iron-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	100.0	30.10	n/a
S96T003978		D Potassium-ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	103.0	6.010	n/a
S96T003978		D Lithium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	< 60.10	<60.10	n/a	n/a	95.00	60.10	n/a
S96T003978		D Magnesium-ICP-Acid Dil.	ug/mL	96.60	<1.00e-01	< 60.10	<60.10	n/a	n/a	94.20	6.010	n/a
S96T003978		D Manganese-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	< 6.010	<60.10	86.90	8.52	99.00	30.10	n/a
S96T003978		D Molybdenum-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	90.60	83.20	86.90	8.35	-7.100e0	60.10	n/a
S96T003978		D Sodium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	2.37e+05	2.18e+05	2.28e+05	n/a	160.0	60.10	n/a
S96T003978		D Neodymium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	< 60.10	<60.10	n/a	n/a	97.10	12.00	n/a
S96T003978		D Nickel-ICP-Acid Dil.	ug/mL	99.80	<2.00e-02	< 12.00	<1.20e1	328.5	15.5	96.00	120.0	n/a
S96T003978		D Phosphorus-ICP-Acid Dil.	ug/mL	100.4	<2.00e-01	3.54e+02	303.0	328.5	n/a	106.0	60.10	n/a
S96T003978		D Lead-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	< 60.10	<60.10	n/a	n/a	103.0	60.10	n/a
S96T003978		D Sulfur-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	9.17e+02	852.0	884.5	7.35	95.40	36.10	n/a
S96T003978		D Antimony-ICP-Acid Dil.	ug/mL	95.80	<6.00e-02	< 36.10	<3.61e1	n/a	n/a	98.40	60.10	n/a
S96T003978		D Selenium-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	< 60.10	<60.10	n/a	n/a	96.20	30.10	n/a
S96T003978		D Silicon-ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	2.28e+02	210.0	219.0	8.22	96.20	60.10	n/a
S96T003978		D Samarium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	< 60.10	<60.10	n/a	n/a	95.80	6.010	n/a
S96T003978		D Strontium-ICP-Acid Dil.	ug/mL	96.60	<1.00e-02	< 6.010	<60.10	n/a	n/a	95.50	6.010	n/a
S96T003978		D Titanium-ICP-Acid Dil.	ug/mL	97.60	<1.00e-02	< 6.010	<60.10	n/a	n/a	94.60	120.0	n/a
S96T003978		D Thallium-ICP-Acid Dil.	ug/mL	96.20	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	95.00	300.0	n/a
S96T003978		D Uranium-ICP-Acid Dil.	ug/mL	96.10	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	98.40	30.10	n/a
S96T003978		D Vanadium-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 30.10	<3.01e1	12.95	6.95	99.90	6.010	n/a
S96T003978		D Zinc-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	13.40	12.50	12.95	n/a	97.90	6.010	n/a
S96T003978		D Zirconium-ICP-Acid Dil.	ug/mL	98.60	<2.00e-02	< 6.010	<60.10	622.0	8.36	n/a	103.0	n/a
S96T003978		D Acetate by IC-Dipnax 4000	ug/mL	109.7	<2.00e-02	5.96e+02	648.0	622.0	n/a	n/a	103.0	n/a
S96T003978		D Formate by Ion Chromatograph	ug/mL	100.0	<2.10e-02	1.59e+03	1.68e+03	1.63e+03	5.50	n/a	108.2	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003978		Fluoride-IC-Dionex 4000/4500	ug/mL	100.7	<1.30e-02	< 53.83	<5.38e1	n/a	n/a	n/a	53.83	n/a
S961003978		Chloride-IC-Dionex 4000/4500	ug/mL	96.08	<1.70e-02	9.46e+03	9.47e+03	9.46e+03	0.11	n/a	70.40	n/a
S961003978		Nitrite-IC - Dionex 4000/4500	ug/mL	95.59	<1.07e-01	1.22e+05	1.22e+05	1.22e+05	0.00	n/a	443.1	n/a
S961003978		Bromide By Ion Chromatograph	ug/mL	99.49	<1.26e-01	7.68e+02	785.0	776.5	2.19	n/a	521.8	n/a
S961003978		Nitrate By IC-Dionex 4000/4500	ug/mL	97.14	<1.50e-01	1.60e+05	1.60e+05	1.60e+05	0.00	n/a	579.7	n/a
S961003978		Phosphate-IC-Dionex 4000/4500	ug/mL	95.23	<1.19e-01	8.38e+02	603.0	720.3	32.6	n/a	492.4	n/a
S961003978		Sulfate By IC-Dionex 4000/4500	ug/mL	97.46	<1.36e-01	1.18e+03	1.08e+03	1.13e+03	8.85	n/a	562.8	n/a
S961003978		Oxalate By IC-Dionex 4000/4500	ug/mL	99.43	<1.05e-01	<4.35e+02	<4.35e2	n/a	n/a	n/a	436.8	n/a
S961003978		Alpha In Liquid Samples	uCi/mL	93.75	<2.94e-02	<2.58e-02	<1.86E-2	n/a	n/a	n/a	3.90e-02	5.00E+02

CORE NUMBER: 153
SEGMENT #: 14Table 1. Data Summary Table
AN-105

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Er%
S96T004334		Bulk Density of Sample	g/mL	n/a	n/a	1.560	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S96T004335		DSC Exotherm using Mettler	Joules/g	93.85	n/a	31.00	39.00	35.00	22.9	n/a	n/a	n/a	n/a
S96T004335		DSC Exotherm Dry Calculated	Joules/g	93.85	n/a	57.80	72.70	65.25	22.8	n/a	n/a	n/a	n/a
S96T004335		% Water by TGA using Mettler	%	98.28	n/a	46.87	45.88	46.38	2.13	n/a	n/a	n/a	n/a
S96T004337	F	Silver - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.03e-02	<2.03e-02	2.66e+04	4.14	90.10	203.0	n/a	n/a
S96T004337	F	Aluminum - ICP-Fusion	ug/g	96.60	<5.00e-02	<2.71e+04	2.60e+04	2.60e+04	4.14	90.10	203.0	n/a	n/a
S96T004337	F	Arsenic - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	102.0	2.03e+03	n/a	n/a
S96T004337	F	Boron - ICP-Fusion	ug/g	97.40	<5.00e-02	<1.02e+03	<1.01e+03	n/a	n/a	95.90	1.02e+03	n/a	n/a
S96T004337	F	Barium - ICP-Fusion	ug/g	97.20	<5.00e-02	<1.02e+03	<1.01e+03	n/a	n/a	95.90	1.02e+03	n/a	n/a
S96T004337	F	Beryllium - ICP-Fusion	ug/g	100.6	<5.00e-03	<1.02e+03	<1.01e+03	n/a	n/a	98.50	102.0	n/a	n/a
S96T004337	F	Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	97.90	2.03e+03	n/a	n/a
S96T004337	F	Calcium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	97.90	2.03e+03	n/a	n/a
S96T004337	F	Cadmium - ICP-Fusion	ug/g	99.60	<5.00e-03	<1.02e+02	<1.01e+02	n/a	n/a	97.40	102.0	n/a	n/a
S96T004337	F	Cerium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	97.40	2.03e+03	n/a	n/a
S96T004337	F	Cobalt - ICP-Fusion	ug/g	99.60	<2.00e-02	<4.06e+02	<4.06e+02	4.06e+02	n/a	97.10	406.0	n/a	n/a
S96T004337	F	Chromium - ICP-Fusion	ug/g	99.80	<1.00e-02	1.93e+03	1.98e+03	1.96e+03	2.56	97.60	203.0	n/a	n/a
S96T004337	F	Copper - ICP-Fusion	ug/g	101.0	<1.00e-02	<1.02e+02	<1.02e+02	n/a	n/a	100.0	203.0	n/a	n/a
S96T004337	F	Iron - ICP-Fusion	ug/g	99.20	<5.00e-02	<1.02e+02	<1.01e+02	n/a	n/a	97.40	1.02e+03	n/a	n/a
S96T004337	F	Potassium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.02e+02	<1.01e+02	n/a	n/a	98.50	1.02e+03	n/a	n/a
S96T004337	F	Lanthanum - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.03e+02	<2.03e+02	n/a	n/a	98.70	203.0	n/a	n/a
S96T004337	F	Lithium - ICP-Fusion	ug/g	92.60	<1.00e-01	<2.03e+03	<2.03e+03	n/a	n/a	95.30	2.03e+03	n/a	n/a
S96T004337	F	Magnesium - ICP-Fusion	ug/g	97.00	<1.00e-02	<2.03e+02	<2.03e+02	n/a	n/a	95.10	203.0	n/a	n/a
S96T004337	F	Manganese - ICP-Fusion	ug/g	99.80	<5.00e-02	<1.02e+03	<1.01e+03	n/a	n/a	98.30	1.02e+03	n/a	n/a
S96T004337	F	Molybdenum - ICP-Fusion	ug/g	96.00	1.580	2.22e+05	2.15e+05	2.18e+05	3.20	54.00	2.03e+03	n/a	n/a
S96T004337	F	Sodium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	98.90	2.03e+03	n/a	n/a
S96T004337	F	Neodymium - ICP-Fusion	ug/g	98.40	<4.00e-02	<7.12e+02	<4.06e+02	n/a	n/a	97.30	406.0	n/a	n/a
S96T004337	F	Nickel - ICP-Fusion	ug/g	102.8	<2.00e-01	<4.06e+03	<4.06e+03	4.06e+03	n/a	104.0	4.06e+03	n/a	n/a
S96T004337	F	Phosphorus - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.03e+03	<2.03e+03	2.03e+03	n/a	96.40	2.03e+03	n/a	n/a
S96T004337	F	Lead - ICP-Fusion	ug/g	98.20	<1.00e-01	<2.03e+02	<2.03e+02	2.08e+03	3.36	97.50	2.03e+03	n/a	n/a
S96T004337	F	Sulfur - ICP-Fusion	ug/g	94.80	<6.00e-02	<1.22e+03	<1.22e+03	n/a	n/a	92.50	1.22e+03	n/a	n/a
S96T004337	F	Antimony - ICP-Fusion	ug/g	95.60	<1.00e-01	<2.03e+03	<2.03e+03	n/a	n/a	96.40	2.03e+03	n/a	n/a
S96T004337	F	Selenium - ICP-Fusion	ug/g	95.60	<1.00e-01	<2.03e+03	<2.03e+03	n/a	n/a	96.40	2.03e+03	n/a	n/a
S96T004337	F	Silicon - ICP-Fusion	ug/g	99.80	<1.3e-01	<1.02e+03	<2.03e+03	n/a	n/a	94.10	2.03e+03	n/a	n/a
S96T004337	F	Samarium - ICP-Fusion	ug/g	96.00	<1.00e-01	<2.03e+03	<2.03e+03	n/a	n/a	95.60	203.0	n/a	n/a
S96T004337	F	Strontium - ICP-Fusion	ug/g	96.80	<1.00e-02	<2.03e+02	<2.03e+02	n/a	n/a	95.70	203.0	n/a	n/a
S96T004337	F	Titanium - ICP-Fusion	ug/g	95.40	<5.00e-01	<4.06e+03	<4.06e+03	n/a	n/a	90.60	4.06e+03	n/a	n/a
S96T004337	F	Thallium - ICP-Fusion	ug/g	94.30	<5.00e-01	<1.02e+04	<1.01e+04	n/a	n/a	89.00	1.02e+04	n/a	n/a
S96T004337	F	Uranium - ICP-Fusion	ug/g	98.80	<1.00e-02	<1.02e+03	<1.01e+03	n/a	n/a	96.90	1.02e+03	n/a	n/a
S96T004337	F	Vanadium - ICP-Fusion	ug/g	100.6	<5.00e-02	<2.03e+02	<2.03e+02	n/a	n/a	99.00	203.0	n/a	n/a
S96T004337	F	Zinc - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.03e+02	<2.03e+02	n/a	n/a	95.60	203.0	n/a	n/a
S96T004337	F	Zirconium - ICP-Fusion	ug/g	107.0	<3.04e-03	<3.84e-03	4.09e-03	n/a	n/a	84.12	5.00e-03	n/a	n/a
S96T004337	F	Alpha of Digested Solid	ug/g	44.70	<1.00e-02	13.80	13.40	13.60	2.94	57.60	4.680	n/a	n/a
S96T004338	A	Silver - ICP-Acid Digest	ug/g	44.70	<1.00e-02	13.80	13.40	13.60	2.94	57.60	4.680	n/a	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004338	A	Aluminum - ICP-Acid Digest	ug/g	98.80	1.22e-01	2.37e+04	2.39e+04	2.38e+04	0.84	43.00	23.40	n/a
S96T004338	A	Arsenic - ICP-Acid Digest	ug/g	97.80	<1.00e-01	<1.45e+01	<4.78e1	n/a	n/a	92.00	46.80	n/a
S96T004338	A	Boron - ICP-Acid Digest	ug/g	106.8	4.15e-01	1.16e+02	77.20	96.10	39.3	90.80	23.40	n/a
S96T004338	A	Barium - ICP-Acid Digest	ug/g	96.00	<5.00e-02	<2.35e+01	<2.39e1	n/a	n/a	90.80	23.40	n/a
S96T004338	A	Beryllium - ICP-Acid Digest	ug/g	99.40	<5.00e-03	<2.35e+01	<2.39e1	n/a	n/a	91.60	2.330	n/a
S96T004338	A	Bismuth - ICP-Acid Digest	ug/g	94.80	<1.00e-01	1.05e+03	<4.78e1	n/a	n/a	89.00	46.80	n/a
S96T004338	A	Calcium - ICP-Acid Digest	ug/g	180.8	4.180	4.67e+03	788.0	810.0	28.5	71.00	46.80	n/a
S96T004338	A	Cadmium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	7.57e+01	8.81e+01	8.190	15.1	89.60	2.330	n/a
S96T004338	A	Cerium - ICP-Acid Digest	ug/g	98.00	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	93.00	46.80	n/a
S96T004338	A	Cobalt - ICP-Acid Digest	ug/g	95.60	<2.00e-02	<9.34e+01	<9.55e1	n/a	n/a	90.80	9.350	n/a
S96T004338	A	Chromium - ICP-Acid Digest	ug/g	94.40	<1.00e-02	1.84e+03	2.01e+03	1.92e+03	8.83	97.60	4.680	n/a
S96T004338	A	Copper - ICP-Acid Digest	ug/g	94.60	<1.00e-02	5.70e+01	6.91e+01	6.305	19.2	88.60	2.340	n/a
S96T004338	A	Iron - ICP-Acid Digest	ug/g	94.80	<5.00e-02	5.46e+01	61.40	58.00	11.7	90.20	23.40	n/a
S96T004338	A	Potassium - ICP-Acid Digest	ug/g	97.80	<5.00e-01	3.78e+03	3.82e+03	3.85e+03	1.83	81.00	234.0	n/a
S96T004338	A	Lanthanum - ICP-Acid Digest	ug/g	98.20	<5.00e-02	<2.35e+01	<2.39e1	n/a	n/a	93.80	4.680	n/a
S96T004338	A	Lithium - ICP-Acid Digest	ug/g	99.20	<1.00e-01	7.91e+01	8.34e+01	8.125	5.29	95.00	46.80	n/a
S96T004338	A	Magnesium - ICP-Acid Digest	ug/g	90.80	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	87.00	46.80	n/a
S96T004338	A	Manganese - ICP-Acid Digest	ug/g	92.20	<1.00e-02	5.300e+01	5.770e+01	5.535	8.49	86.00	4.680	n/a
S96T004338	A	Molybdenum - ICP-Acid Digest	ug/g	93.00	<5.00e-02	53.50e+01	51.80e+01	52.65	3.23	87.80	23.40	n/a
S96T004338	A	Sodium - ICP-Acid Digest	ug/g	134.4	8.91e-01	1.72e+05	1.74e+05	1.73e+05	1.16	-1.100e2	46.80	n/a
S96T004338	A	Neodymium - ICP-Acid Digest	ug/g	97.00	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	92.40	46.80	n/a
S96T004338	A	Nickel - ICP-Acid Digest	ug/g	96.80	<2.00e-02	28.70e+01	34.00e+01	31.35	16.9	91.00	9.350	n/a
S96T004338	A	Phosphorus - ICP-Acid Digest	ug/g	98.80	<2.00e-01	4.92e+02	485.0	488.5	1.43	91.80	93.50	n/a
S96T004338	A	Lead - ICP-Acid Digest	ug/g	91.60	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	92.00	46.80	n/a
S96T004338	A	Sulfur - ICP-Acid Digest	ug/g	96.60	1.10e-01	2.05e+03	2.13e+03	2.13e+03	7.51	90.80	46.80	n/a
S96T004338	A	Antimony - ICP-Acid Digest	ug/g	94.00	<6.00e-02	<28.00e+01	<2.87e1	n/a	n/a	89.80	28.00	n/a
S96T004338	A	Selenium - ICP-Acid Digest	ug/g	90.20	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	85.20	46.80	n/a
S96T004338	A	Silicon - ICP-Acid Digest	ug/g	118.4	3.48e-01	1.42e+02	<2.39e1	n/a	n/a	91.40	23.40	n/a
S96T004338	A	Samarium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	<4.67e+01	<4.78e1	n/a	n/a	90.20	46.80	n/a
S96T004338	A	Strontium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	<4.67e+01	<4.78e1	n/a	n/a	90.40	4.680	n/a
S96T004338	A	Titanium - ICP-Acid Digest	ug/g	87.80	<1.00e-02	<4.67e+01	<4.78e1	n/a	n/a	83.00	4.680	n/a
S96T004338	A	Thallium - ICP-Acid Digest	ug/g	89.60	<2.00e-01	<93.40e+01	<9.55e1	n/a	n/a	84.40	93.50	n/a
S96T004338	A	Uranium - ICP-Acid Digest	ug/g	92.80	<5.00e-01	<2.35e+02	<2.39e2	n/a	n/a	90.00	234.0	n/a
S96T004338	A	Vanadium - ICP-Acid Digest	ug/g	93.60	<5.00e-02	<2.35e+01	<2.39e1	n/a	n/a	88.40	23.40	n/a
S96T004338	A	Zinc - ICP-Acid Digest	ug/g	91.80	1.03e-01	57.90e+01	55.90e+01	55.95	6.97	87.60	4.680	n/a
S96T004338	A	Zirconium - ICP-Acid Digest	ug/g	96.20	<1.00e-02	4.920e+01	4.850e+01	4.885	1.43	89.00	46.80	n/a
S96T004339	M	Acetate by IC-Dionex 4000/4500	ug/g	100.8	<2.00e-02	8.27e+02	808.0	817.5	2.32	119.4	79.80	n/a
S96T004339	M	Formate by IC-Dionex 4000/4500	ug/g	102.5	<1.00e-02	1.17e+03	1.18e+03	1.18e+03	0.85	109.2	83.79	n/a
S96T004339	M	Fluoride by IC-Dionex 4000/4500	ug/g	100.7	<1.00e-02	1.41e+03	1.51e+03	1.28e+03	20.3	110.7	51.87	n/a
S96T004339	M	Chloride by IC-Dionex 4000/4500	ug/g	102.2	2.30e-02	7.25e+03	6.05e+03	6.65e+03	18.0	60.00	67.83	n/a
S96T004339	M	Nitrite by IC-Dionex 4000/4500	ug/g	100.9	1.98e-01	8.72e+04	8.24e+04	8.24e+04	11.8	47.88	426.9	n/a
S96T004339	M	Bromide by Ion Chromatograph	ug/g	101.2	<1.26e-01	5.06e+02	<5.03e2	n/a	n/a	94.57	502.7	n/a
S96T004339	M	Nitrate by IC-Dionex 4000/4500	ug/g	100.3	1.76e-01	1.18e+05	1.02e+05	1.10e+05	14.5	23.57	558.2	n/a
S96T004339	M	Phosphate by IC-Dionex 4000/4500	ug/g	100.9	<1.19e-01	1.20e+03	1.34e+03	1.27e+03	11.0	98.35	474.4	n/a
S96T004339	M	Sulfate by IC-Dionex 4000/4500	ug/g	102.2	<1.36e-01	6.94e+03	6.15e+03	6.51e+03	25.7	97.31	542.2	n/a
S96T004339	M	Oxalate by IC-Dionex 4000/4500	ug/g	100.6	<1.05e-01	1.17e+04	8.58e+03	1.01e+04	30.8	92.63	418.9	n/a

Drainable Liquid: Drainable Liquid

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
5961004333		DSC Exotherm Dry. Calculated	Joules/g Dry	n/a	n/a	69.10	115.0	92.05	49.9	n/a	n/a	n/a	n/a
5961004333		DSC Exotherm on Perkin Elmer	Joules/g	96.59	n/a	34.80	57.70	46.25	49.5	n/a	n/a	n/a	n/a
5961004333		Specific Gravity	Sp.G.	97.00	n/a	1.450	1.445	1.448	0.35	n/a	1.00e-02	n/a	n/a
5961004333		% Water by TGA on Perkin Elmer	%	99.09	n/a	49.65	49.65	49.63	0.06	n/a	n/a	n/a	n/a
5961004333		D Silver-ICP-Acid Dil.	ug/mL	101.0	<1.00e-02	19.20	19.40	19.30	1.04	92.40	6.010	n/a	n/a
5961004333		D Arsenic-ICP-Acid Dil.	ug/mL	98.60	<5.00e-02	4.43e+04	4.44e+04	4.44e+04	0.23	179.0	30.10	n/a	n/a
5961004333		D Boron-ICP-Acid Dil.	ug/mL	104.0	<1.00e-01	60.10	<6.01e1	n/a	n/a	97.80	60.10	n/a	n/a
5961004333		D Barium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	77.80	74.60	76.20	4.20	102.0	30.10	n/a	n/a
5961004333		D Beryllium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	3.000	<3.00e0	n/a	n/a	99.50	30.10	n/a	n/a
5961004333		D Bismuth-ICP-Acid Dil.	ug/mL	102.6	<5.00e-01	60.10	<6.01e1	n/a	n/a	103.0	3.000	n/a	n/a
5961004333		D Calcium-ICP-Acid Dil.	ug/mL	102.2	<1.00e-01	60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a	n/a
5961004333		D Cadmium-ICP-Acid Dil.	ug/mL	102.0	<5.00e-03	3.000	<3.00e0	n/a	n/a	104.0	3.000	n/a	n/a
5961004333		D Cerium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	60.10	<6.01e1	n/a	n/a	102.0	60.10	n/a	n/a
5961004333		D Cobalt-ICP-Acid Dil.	ug/mL	103.4	<2.00e-02	12.00	<1.20e1	n/a	n/a	102.0	12.00	n/a	n/a
5961004333		D Chromium-ICP-Acid Dil.	ug/mL	102.6	<1.00e-02	2.16e+02	220.0	218.0	1.83	102.0	6.010	n/a	n/a
5961004333		D Copper-ICP-Acid Dil.	ug/mL	104.8	<1.00e-02	6.010	<6.01e0	n/a	n/a	105.0	6.010	n/a	n/a
5961004333		D Iron-ICP-Acid Dil.	ug/mL	101.8	<5.00e-01	30.10	<3.01e1	n/a	n/a	101.0	30.10	n/a	n/a
5961004333		D Potassium-ICP-Acid Dil.	ug/mL	101.8	<5.00e-01	7.14e+03	7.05e+03	7.10e+03	1.27	99.70	300.0	n/a	n/a
5961004333		D Lanthanum-ICP-Acid Dil.	ug/mL	101.6	<5.00e-02	30.10	<3.01e1	n/a	n/a	103.0	30.10	n/a	n/a
5961004333		D Lithium-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	6.010	<6.01e0	n/a	n/a	106.0	6.010	n/a	n/a
5961004333		D Magnesium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	60.10	<6.01e1	n/a	n/a	95.20	60.10	n/a	n/a
5961004333		D Manganese-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	6.010	<6.01e0	n/a	n/a	96.80	6.010	n/a	n/a
5961004333		D Molybdenum-ICP-Acid Dil.	ug/mL	103.8	<5.00e-02	1.03e+02	103.0	103.0	0.00	104.0	30.10	n/a	n/a
5961004333		D Sodium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	2.63e+05	2.63e+05	2.64e+05	1.13	-3.66e02	60.10	n/a	n/a
5961004333		D Neodymium-ICP-Acid Dil.	ug/mL	102.0	<1.00e-01	60.10	<6.01e1	n/a	n/a	103.0	60.10	n/a	n/a
5961004333		D Nickel-ICP-Acid Dil.	ug/mL	102.4	<2.00e-02	12.00	<1.20e1	n/a	n/a	103.0	12.00	n/a	n/a
5961004333		D Phosphorus-ICP-Acid Dil.	ug/mL	104.0	<2.00e-01	3.74e+02	371.0	372.5	0.81	103.0	120.0	n/a	n/a
5961004333		D Lead-ICP-Acid Dil.	ug/mL	102.4	<1.00e-01	60.10	<6.01e1	n/a	n/a	109.0	60.10	n/a	n/a
5961004333		D Sulfur-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	9.77e+02	1.02e+03	998.5	4.31	119.0	36.10	n/a	n/a
5961004333		D Antimony-ICP-Acid Dil.	ug/mL	97.60	<6.00e-02	36.10	<3.61e1	n/a	n/a	99.20	36.10	n/a	n/a
5961004333		D Selenium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	60.10	<6.01e1	n/a	n/a	100.0	60.10	n/a	n/a
5961004333		D Silicon-ICP-Acid Dil.	ug/mL	97.20	<5.00e-02	1.65e+02	160.0	162.5	3.08	97.70	30.10	n/a	n/a
5961004333		D Samarium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	60.10	<6.01e1	n/a	n/a	99.10	60.10	n/a	n/a
5961004333		D Strontium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	60.10	<6.01e0	n/a	n/a	100.0	6.010	n/a	n/a
5961004333		D Titanium-ICP-Acid Dil.	ug/mL	99.60	<2.00e-02	6.010	<6.01e0	n/a	n/a	99.60	6.010	n/a	n/a
5961004333		D Thallium-ICP-Acid Dil.	ug/mL	97.80	<2.00e-01	<1.20e02	<1.20e2	n/a	n/a	95.60	120.0	n/a	n/a
5961004333		D Uranium-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	<3.00e+02	<3.00e2	n/a	n/a	98.50	300.0	n/a	n/a
5961004333		D Vanadium-ICP-Acid Dil.	ug/mL	97.50	<5.00e-02	30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a	n/a
5961004333		D Zinc-ICP-Acid Dil.	ug/mL	104.4	<1.00e-02	43.40	<42.20	42.80	2.80	104.0	6.010	n/a	n/a
5961004333		D Zirconium-ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	6.010	<6.01e0	n/a	n/a	103.0	6.010	n/a	n/a
5961004333		Acetate by IC-Dionex 4000/4500	ug/mL	104.0	<2.00e-02	7.72e+02	790.0	781.1	2.30	n/a	103.0	n/a	n/a
5961004333		Formate by Ion Chromatograph	ug/mL	105.8	<2.10e-02	1.94e+03	1.93e+03	1.94e+03	0.52	n/a	108.2	n/a	n/a
5961004333		Fluoride-IC-Dionex 4000/4500	ug/mL	102.0	<1.30e-02	4.52e+02	439.0	445.8	2.92	37.97	53.83	n/a	n/a
5961004333		Chloride-IC-Dionex 4000/4500	ug/mL	101.5	<1.70e-02	1.02e+04	1.01e+04	1.01e+04	0.99	162.0	70.40	n/a	n/a
5961004333		Nitrite-IC - Dionex 4000/4500	ug/mL	99.45	<1.07e-01	1.27e+05	1.26e+05	1.26e+05	0.79	88.58	443.1	n/a	n/a
5961004333		Bromide by Ion Chromatograph	ug/mL	100.8	<1.26e-01	8.29e+02	829.0	829.0	0.00	94.40	521.8	n/a	n/a
5961004333		Nitrate by IC-Dionex 4000/4500	ug/mL	101.5	1.40e-01	1.73e+05	1.73e+05	1.73e+05	0.00	87.21	579.7	n/a	n/a

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Sample#	R/A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S96T004333		Phosphate-IC-Dionex 4000/4500	ug/mL	99.63	<1.19e-01	6.40e-02	609.0	634.5	8.04	97.98	492.4		n/a
S96T004333		Sulfate by IC-Dionex 4000/4500	ug/mL	101.6	<1.36e-01	5.21e-03	1.37e+03	1.29e+03	12.4	102.1	562.8		n/a
S96T004333		Oxalate by IC-Dionex 4000/4500	ug/mL	100.4	<1.05e-01	5.28e-02	682.0	625.2	18.2	98.68	434.8		n/a
S96T004333		Alpha in Liquid Samples	uCi/mL	95.31	1.10e-02	<7.10e-03	1.40e-02	n/a	n/a	n/a	1.80e-02	5.00E+02	

CORE NUMBER: 153
SEGMENT #: 15

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961004302		Bulk Density of Sample	g/mL	n/a	n/a	1.520	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961004310		DSC Exotherm using Mettler	Joules/g	105.1	n/a	40.40	23.90	32.15	51.3	n/a	n/a	n/a
S961004310		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	77.40	45.80	61.60	51.3	n/a	n/a	n/a
S961004310		% Water by TGA using Mettler	%	98.31	n/a	47.55	48.04	47.80	1.03	n/a	n/a	n/a
S961004314		F Silver - ICP-Fusion	ug/g	98.40	<1.00e-02	<1.91e+02	<1.92e2	2.47e+04	n/a	90.90	191.0	n/a
S961004314		F Aluminum - ICP-Fusion	ug/g	96.60	<5.00e-02	2.49e+04	2.45e+04	2.47e+04	1.62	90.50	951.0	n/a
S961004314		F Arsenic - ICP-Fusion	ug/g	101.2	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	101.0	1.91e+03	n/a
S961004314		F Barium - ICP-Fusion	ug/g	97.40	<5.00e-02	<9.58e+02	<9.58e2	n/a	n/a	95.30	951.0	n/a
S961004314		F Beryllium - ICP-Fusion	ug/g	97.20	<5.00e-02	<9.53e+02	<9.58e2	n/a	n/a	98.60	95.10	n/a
S961004314		F Bismuth - ICP-Fusion	ug/g	100.6	<5.00e-03	<9.30	<9.58e1	n/a	n/a	95.20	1.91e+03	n/a
S961004314		F Calcium - ICP-Fusion	ug/g	100.0	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	98.10	1.91e+03	n/a
S961004314		F Cadmium - ICP-Fusion	ug/g	99.60	<5.00e-03	<9.30	<9.58e1	n/a	n/a	97.30	95.10	n/a
S961004314		F Cerium - ICP-Fusion	ug/g	98.00	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	97.40	1.91e+03	n/a
S961004314		F Cobalt - ICP-Fusion	ug/g	99.60	<2.00e-02	<3.81e+02	<3.83e2	1.42e+03	13.3	97.40	381.0	n/a
S961004314		F Chromium - ICP-Fusion	ug/g	99.80	<1.00e-02	1.33e+03	1.52e+03	1.42e+03	13.3	97.40	191.0	n/a
S961004314		F Copper - ICP-Fusion	ug/g	101.0	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	100.0	191.0	n/a
S961004314		F Iron - ICP-Fusion	ug/g	99.20	<5.00e-02	<9.53e+02	<9.58e2	n/a	n/a	97.20	951.0	n/a
S961004314		F Potassium - ICP-Fusion	ug/g	99.00	<5.00e-01	5.70e+06	n/a	n/a	n/a	n/a	9.51e+03	n/a
S961004314		F Lanthanum - ICP-Fusion	ug/g	99.60	<5.00e-02	<9.53e+02	<9.58e2	n/a	n/a	98.70	951.0	n/a
S961004314		F Lithium - ICP-Fusion	ug/g	98.80	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	98.60	191.0	n/a
S961004314		F Magnesium - ICP-Fusion	ug/g	92.60	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	93.10	1.91e+03	n/a
S961004314		F Manganese - ICP-Fusion	ug/g	97.00	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	93.10	191.0	n/a
S961004314		F Molybdenum - ICP-Fusion	ug/g	99.80	<5.00e-02	<9.53e+02	<9.58e2	n/a	n/a	98.50	951.0	n/a
S961004314		F Sodium - ICP-Fusion	ug/g	96.00	1.580	2.15e+05	2.17e+05	2.16e+05	0.93	65.00	1.91e+03	n/a
S961004314		F Neodymium - ICP-Fusion	ug/g	99.40	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	98.50	1.91e+03	n/a
S961004314		F Nickel - ICP-Fusion	ug/g	102.8	4.50e-02	7.67e+02	543.0	655.0	34.2	95.80	381.0	n/a
S961004314		F Phosphorus - ICP-Fusion	ug/g	99.00	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	107.0	3.81e+03	n/a
S961004314		F Lead - ICP-Fusion	ug/g	98.20	<1.00e-01	2.06e+03	2.55e+03	2.30e+03	21.3	97.40	1.91e+03	n/a
S961004314		F Sulfur - ICP-Fusion	ug/g	94.80	<6.00e-02	<1.14e+03	<1.15e3	n/a	n/a	93.80	1.14e+03	n/a
S961004314		F Antimony - ICP-Fusion	ug/g	95.60	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	94.50	1.91e+03	n/a
S961004314		F Selenium - ICP-Fusion	ug/g	99.80	1.13e-01	<9.53e+02	<9.58e2	n/a	n/a	101.0	951.0	n/a
S961004314		F Silicon - ICP-Fusion	ug/g	96.00	<1.00e-01	<1.91e+03	<1.92e3	n/a	n/a	94.20	1.91e+03	n/a
S961004314		F Samarium - ICP-Fusion	ug/g	96.60	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	95.60	191.0	n/a
S961004314		F Strontium - ICP-Fusion	ug/g	96.60	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	96.00	191.0	n/a
S961004314		F Tantalum - ICP-Fusion	ug/g	95.40	<2.00e-01	<3.81e+03	<3.83e3	n/a	n/a	93.50	3.81e+03	n/a
S961004314		F Thallium - ICP-Fusion	ug/g	94.30	<5.00e-01	<9.53e+03	<9.58e3	n/a	n/a	89.00	9.51e+03	n/a
S961004314		F Uranium - ICP-Fusion	ug/g	98.80	<5.00e-02	<9.53e+02	<9.58e2	n/a	n/a	96.90	951.0	n/a
S961004314		F Zinc - ICP-Fusion	ug/g	100.6	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	95.70	191.0	n/a
S961004314		F Zirconium - ICP-Fusion	ug/g	97.60	<1.00e-02	<1.91e+02	<1.92e2	n/a	n/a	95.70	191.0	n/a
S961004315		A Silver - ICP-Acid Digest	ug/g	44.70	<1.00e-02	12.20	12.10	12.15	0.82	n/a	4.770	n/a
S961004315		A Aluminum - ICP-Acid Digest	ug/g	98.80	1.22e-01	2.31e+04	2.23e+04	2.27e+04	3.52	n/a	23.80	n/a

R	Sample#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
	S961004315	A Arsenic - ICP-Acid Digest	ug/g	97.80	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Boron - ICP-Acid Digest	ug/g	106.8	4.15e-01	1.03e+02	84.50	93.75	19.7	N/A	23.80	N/A	N/A
	S961004315	A Barium - ICP-Acid Digest	ug/g	96.00	<5.00e-02	< 23.80	<2.39e1	N/A	N/A	N/A	23.80	N/A	N/A
	S961004315	A Beryllium - ICP-Acid Digest	ug/g	99.40	<1.00e-03	< 2.380	<2.39e0	N/A	N/A	N/A	2.380	N/A	N/A
	S961004315	A Bismuth - ICP-Acid Digest	ug/g	94.80	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Calcium - ICP-Acid Digest	ug/g	180.8	4.180	7.08e+02	856.0	782.0	18.9	N/A	47.80	N/A	N/A
	S961004315	A Cadmium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	< 5.490	<5.240	5.365	4.66	N/A	2.380	N/A	N/A
	S961004315	A Cerium - ICP-Acid Digest	ug/g	98.00	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Cobalt - ICP-Acid Digest	ug/g	95.60	<2.00e-02	< 9.530	<9.56e0	N/A	N/A	N/A	9.530	N/A	N/A
	S961004315	A Chromium - ICP-Acid Digest	ug/g	94.40	<1.00e-02	1.35e+03	1.40e+03	1.38e+03	3.64	N/A	4.770	N/A	N/A
	S961004315	A Copper - ICP-Acid Digest	ug/g	94.60	<1.00e-02	8.100	5.850	6.975	32.3	N/A	23.80	N/A	N/A
	S961004315	A Iron - ICP-Acid Digest	ug/g	94.80	<5.00e-02	83.60	62.50	73.05	28.9	N/A	23.80	N/A	N/A
	S961004315	A Potassium - ICP-Acid Digest	ug/g	97.80	<5.00e-01	3.56e+03	3.50e+03	3.53e+03	1.70	N/A	23.80	N/A	N/A
	S961004315	A Lanthanum - ICP-Acid Digest	ug/g	98.20	<1.00e-02	< 23.80	<2.39e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Lithium - ICP-Acid Digest	ug/g	99.20	<1.00e-02	< 29.00	<30.20	29.60	4.05	N/A	47.80	N/A	N/A
	S961004315	A Magnesium - ICP-Acid Digest	ug/g	90.80	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Manganese - ICP-Acid Digest	ug/g	92.20	<1.00e-02	< 5.150	<4.860	5.005	5.79	N/A	47.80	N/A	N/A
	S961004315	A Molybdenum - ICP-Acid Digest	ug/g	134.4	<5.00e-02	53.30	49.30	51.30	7.80	N/A	23.80	N/A	N/A
	S961004315	A Sodium - ICP-Acid Digest	ug/g	97.00	<1.00e-01	< 47.60	<4.78e1	1.68e+05	0.59	N/A	47.80	N/A	N/A
	S961004315	A Neodymium - ICP-Acid Digest	ug/g	96.80	<2.00e-02	< 23.70	<24.30	24.00	2.50	N/A	9.530	N/A	N/A
	S961004315	A Nickel - ICP-Acid Digest	ug/g	98.80	<2.00e-01	3.60e+02	380.0	370.0	5.41	N/A	95.20	N/A	N/A
	S961004315	A Phosphorus - ICP-Acid Digest	ug/g	91.60	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Lead - ICP-Acid Digest	ug/g	96.60	<1.00e-01	2.19e+03	2.27e+03	2.23e+03	3.59	N/A	47.80	N/A	N/A
	S961004315	A Sulfur - ICP-Acid Digest	ug/g	94.00	<1.00e-02	< 28.60	<2.87e1	N/A	N/A	N/A	28.50	N/A	N/A
	S961004315	A Antimony - ICP-Acid Digest	ug/g	90.20	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Selenium - ICP-Acid Digest	ug/g	118.4	<3.48e-01	1.75e+02	180.0	177.5	2.82	N/A	23.80	N/A	N/A
	S961004315	A Silicon - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 47.60	<4.78e1	N/A	N/A	N/A	47.80	N/A	N/A
	S961004315	A Samarium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	< 4.760	<4.78e0	N/A	N/A	N/A	4.770	N/A	N/A
	S961004315	A Strontium - ICP-Acid Digest	ug/g	87.80	<1.00e-02	< 4.760	<4.78e0	N/A	N/A	N/A	4.770	N/A	N/A
	S961004315	A Titanium - ICP-Acid Digest	ug/g	89.60	<2.00e-01	< 95.30	<9.56e1	N/A	N/A	N/A	95.20	N/A	N/A
	S961004315	A Thallium - ICP-Acid Digest	ug/g	92.80	<5.00e-01	<2.38e+02	<2.39e2	N/A	N/A	N/A	238.0	N/A	N/A
	S961004315	A Uranium - ICP-Acid Digest	ug/g	93.60	<5.00e-02	< 23.80	<2.39e1	N/A	N/A	N/A	23.80	N/A	N/A
	S961004315	A Vanadium - ICP-Acid Digest	ug/g	91.80	<1.00e-01	57.20	42.10	49.65	30.4	N/A	4.770	N/A	N/A
	S961004315	A Zinc - ICP-Acid Digest	ug/g	96.20	<1.00e-02	5.320	5.660	5.490	6.19	N/A	4.770	N/A	N/A
	S961004316	W Acetate by IC-DIONEX 4000/4500	ug/g	100.8	<2.00e-02	5.57e+02	560.0	558.6	0.54	N/A	78.29	N/A	N/A
	S961004316	W Formate by IC-DIONEX 4000/4500	ug/g	102.5	<2.00e-02	1.19e+03	1.14e+03	1.17e+03	4.29	N/A	82.19	N/A	N/A
	S961004316	W Fluoride by IC-DIONEX 4000/4500	ug/g	100.7	<1.30e-02	1.13e+03	1.17e+03	1.15e+03	3.48	N/A	50.88	N/A	N/A
	S961004316	W Chloride by IC-DIONEX 4000/4500	ug/g	102.2	<2.30e-02	5.85e+03	5.53e+03	5.69e+03	5.62	N/A	66.55	N/A	N/A
	S961004316	W Nitrite by IC-DIONEX 4000/4500	ug/g	100.9	<1.98e-01	7.28e+02	7.00e+02	7.14e+02	3.92	N/A	418.7	N/A	N/A
	S961004316	W Bromide by Ion Chromatography	ug/g	101.2	<1.20e-01	7.58e+02	730.0	744.1	3.76	N/A	493.3	N/A	N/A
	S961004316	W Nitrate by IC-DIONEX 4000/4500	ug/g	100.3	<1.76e-01	1.02e+05	1.03e+05	1.03e+05	0.98	N/A	547.7	N/A	N/A
	S961004316	W Phosphate by IC-DIONEX 4000/4500	ug/g	100.9	<1.19e-01	6.79e+02	1.05e+03	862.6	43.5	N/A	465.4	N/A	N/A
	S961004316	W Sulfate by IC-DIONEX 4000/4500	ug/g	102.2	<1.39e-01	5.89e+03	5.89e+03	5.82e+03	2.41	N/A	531.9	N/A	N/A
	S961004316	W Oxalate by IC-DIONEX 4000/4500	ug/g	100.6	<1.05e-01	8.71e+03	8.48e+03	8.59e+03	2.68	N/A	411.0	N/A	N/A

13-Jan-1997 11:37:22
A-0002-3

L Lower Half of Segment: L Lower Half of Segment

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S96T004303		Bulk Density of Sample	g/mL	n/a	n/a	1.560	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S96T004311		DSC Exotherm using Mettler	Joules/g	105.1	n/a	34.70	57.10	45.90	48.8	n/a	n/a	n/a	n/a
S96T004311		DSC Exotherm Dry Calculated	Joules/g	n/a	n/a	71.50	71.50	94.75	49.1	n/a	n/a	n/a	n/a
S96T004311		% Water by TGA using Mettler	%	98.31	n/a	52.82	50.12	51.47	5.25	n/a	n/a	n/a	n/a
S96T004311	F	Silver - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.00e+02	<2.03e2	89.80	10.0	88.30	200.0	n/a	n/a
S96T004317	F	Aluminum - ICP-Fusion	ug/g	96.60	<5.00e-02	2.51e+04	2.27e+04	2.39e+04	10.0	89.80	996.0	n/a	n/a
S96T004317	F	Arsenic - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	101.0	2.00e+03	n/a	n/a
S96T004317	F	Boron - ICP-Fusion	ug/g	97.40	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	96.80	996.0	n/a	n/a
S96T004317	F	Barium - ICP-Fusion	ug/g	97.20	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	98.00	99.60	n/a	n/a
S96T004317	F	Beryllium - ICP-Fusion	ug/g	100.6	<5.00e-03	<9.98e+02	<1.02e3	n/a	n/a	94.20	2.00e+03	n/a	n/a
S96T004317	F	Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	94.60	2.00e+03	n/a	n/a
S96T004317	F	Calcium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	97.70	2.00e+03	n/a	n/a
S96T004317	F	Cadmium - ICP-Fusion	ug/g	99.60	<5.00e-03	<9.98e+02	<1.02e3	n/a	n/a	96.90	99.60	n/a	n/a
S96T004317	F	Cerium - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	96.70	2.00e+03	n/a	n/a
S96T004317	F	Cobalt - ICP-Fusion	ug/g	99.80	<1.00e-02	<3.99e+02	<4.06e2	n/a	n/a	96.70	399.0	n/a	n/a
S96T004317	F	Chromium - ICP-Fusion	ug/g	101.0	<1.00e-02	1.51e+03	1.79e+03	1.65e+03	17.0	97.10	200.0	n/a	n/a
S96T004317	F	Copper - ICP-Fusion	ug/g	99.20	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	96.70	996.0	n/a	n/a
S96T004317	F	Iron - ICP-Fusion	ug/g	99.00	<5.00e-01	5.71e+06	n/a	n/a	n/a	n/a	9.96e+03	n/a	n/a
S96T004317	F	Potassium - ICP-Fusion	ug/g	99.40	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	97.50	996.0	n/a	n/a
S96T004317	F	Lanthanum - ICP-Fusion	ug/g	99.60	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	97.40	200.0	n/a	n/a
S96T004317	F	Lithium - ICP-Fusion	ug/g	98.80	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	92.60	2.00e+03	n/a	n/a
S96T004317	F	Magnesium - ICP-Fusion	ug/g	92.60	<1.00e-01	<2.00e+02	<2.03e2	n/a	n/a	92.60	2.00e+03	n/a	n/a
S96T004317	F	Manganese - ICP-Fusion	ug/g	97.00	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	92.60	200.0	n/a	n/a
S96T004317	F	Molybdenum - ICP-Fusion	ug/g	99.80	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	97.90	996.0	n/a	n/a
S96T004317	F	Sodium - ICP-Fusion	ug/g	96.00	1.580	2.23e+05	2.30e+05	2.26e+05	3.09	45.60	2.00e+03	n/a	n/a
S96T004317	F	Neodymium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	97.50	2.00e+03	n/a	n/a
S96T004317	F	Nickel - ICP-Fusion	ug/g	98.40	4.50e-02	<3.99e+02	<4.06e2	n/a	n/a	97.80	399.0	n/a	n/a
S96T004317	F	Phosphorus - ICP-Fusion	ug/g	102.8	<2.00e-01	<3.99e+03	<4.06e3	n/a	n/a	105.0	3.99e+03	n/a	n/a
S96T004317	F	Lead - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	96.60	2.00e+03	n/a	n/a
S96T004317	F	Sulfur - ICP-Fusion	ug/g	99.00	<1.00e-01	2.39e+03	2.64e+03	2.52e+03	9.94	98.40	2.00e+03	n/a	n/a
S96T004317	F	Antimony - ICP-Fusion	ug/g	94.80	<6.00e-02	<1.20e+03	<1.22e3	n/a	n/a	94.70	2.00e+03	n/a	n/a
S96T004317	F	Selenium - ICP-Fusion	ug/g	95.60	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	102.0	996.0	n/a	n/a
S96T004317	F	Samarium - ICP-Fusion	ug/g	99.80	1.13e-01	<9.98e+02	<1.02e3	n/a	n/a	94.70	2.00e+03	n/a	n/a
S96T004317	F	Strontium - ICP-Fusion	ug/g	96.80	<1.00e-01	<2.00e+02	<2.03e2	n/a	n/a	93.40	2.00e+03	n/a	n/a
S96T004317	F	Tantalum - ICP-Fusion	ug/g	96.60	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	94.30	200.0	n/a	n/a
S96T004317	F	Titanium - ICP-Fusion	ug/g	95.40	<2.00e-01	<3.99e+03	<4.06e3	n/a	n/a	94.80	200.0	n/a	n/a
S96T004317	F	Uranium - ICP-Fusion	ug/g	94.30	<5.00e-01	<9.98e+02	<1.02e3	n/a	n/a	91.10	3.99e+03	n/a	n/a
S96T004317	F	Vanadium - ICP-Fusion	ug/g	98.80	<5.00e-02	<9.98e+02	<1.02e3	n/a	n/a	88.50	9.96e+03	n/a	n/a
S96T004317	F	Zinc - ICP-Fusion	ug/g	100.6	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	96.20	996.0	n/a	n/a
S96T004317	F	Zirconium - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	98.50	200.0	n/a	n/a
S96T004317	F	Alpha of Digested Solid	ug/g	107.0	<3.04e-03	<3.77e-03	5.35e-03	n/a	n/a	94.70	200.0	n/a	n/a
S96T004318	A	Silver - ICP-Acid Digest	ug/g	44.70	<1.00e-02	12.30	12.60	12.60	n/a	84.68	5.00e-03	1.67e+02	n/a
S96T004318	A	Aluminum - ICP-Acid Digest	ug/g	98.80	1.22e-01	2.33e+04	2.32e+04	2.32e+04	0.43	184.4	4.800	n/a	n/a
S96T004318	A	Arsenic - ICP-Acid Digest	ug/g	97.80	<1.00e-01	<47.90	<4.51e1	n/a	n/a	91.40	48.00	n/a	n/a
S96T004318	A	Boron - ICP-Acid Digest	ug/g	106.8	4.13e-01	1.04e+02	91.20	97.60	13.1	88.00	23.90	n/a	n/a
S96T004318	A	Barium - ICP-Acid Digest	ug/g	96.00	<5.00e-02	<24.00	<2.25e1	n/a	n/a	89.00	23.90	n/a	n/a
S96T004318	A	Beryllium - ICP-Acid Digest	ug/g	99.40	<5.00e-03	<2.400	<2.25e0	n/a	n/a	90.00	2.400	n/a	n/a

HNF-SD-WM-DP-199, REV. 1
WHC-SD-WM-DP-199, REV. 1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004318	A	Bismuth - ICP-Acid Digest	ug/g	94.80	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	89.40	48.00	n/a
S96T004318	A	Calcium - ICP-Acid Digest	ug/g	180.8	4.180	5.89e+02	5.040	564.5	8.68	136.4	48.00	n/a
S96T004318	A	Cadmium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	5.040	5.060	5.050	0.40	88.60	2.400	n/a
S96T004318	A	Cerium - ICP-Acid Digest	ug/g	98.00	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	91.00	48.00	n/a
S96T004318	A	Cobalt - ICP-Acid Digest	ug/g	95.60	<2.00e-02	< 9.580	<9.02e0	n/a	n/a	90.00	9.570	n/a
S96T004318	A	Chromium - ICP-Acid Digest	ug/g	94.40	<1.00e-02	1.21e+03	1.33e+03	1.27e+03	9.45	110.4	4.800	n/a
S96T004318	A	Copper - ICP-Acid Digest	ug/g	94.60	<1.00e-02	5.710	6.050	5.880	5.78	87.00	4.800	n/a
S96T004318	A	Iron - ICP-Acid Digest	ug/g	94.80	<5.00e-02	35.80	36.60	36.20	2.21	88.40	24.00	n/a
S96T004318	A	Potassium - ICP-Acid Digest	ug/g	97.80	<5.00e-01	3.72e+03	3.76e+03	3.74e+03	1.07	106.4	23.90	n/a
S96T004318	A	Lanthanum - ICP-Acid Digest	ug/g	98.20	<5.00e-02	< 24.00	<2.25e1	n/a	n/a	92.00	23.90	n/a
S96T004318	A	Lithium - ICP-Acid Digest	ug/g	99.00	<1.00e-02	8.110	8.810	8.460	8.27	93.40	4.800	n/a
S96T004318	A	Magnesium - ICP-Acid Digest	ug/g	90.80	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	85.40	48.00	n/a
S96T004318	A	Manganese - ICP-Acid Digest	ug/g	92.20	<5.00e-02	< 4.790	5.030	n/a	n/a	86.00	4.800	n/a
S96T004318	A	Molybdenum - ICP-Acid Digest	ug/g	93.00	<5.00e-02	50.20	51.40	50.80	2.36	89.40	23.90	n/a
S96T004318	A	Sodium - ICP-Acid Digest	ug/g	134.4	8.91e-01	1.69e+05	1.70e+05	1.70e+05	0.59	1.40e+03	48.00	n/a
S96T004318	A	Neodymium - ICP-Acid Digest	ug/g	97.00	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	91.00	48.00	n/a
S96T004318	A	Nickel - ICP-Acid Digest	ug/g	96.80	<2.00e-02	21.60	19.50	20.55	10.2	90.20	9.570	n/a
S96T004318	A	Phosphorus - ICP-Acid Digest	ug/g	98.80	<1.00e-01	3.99e+02	372.0	385.5	7.00	94.00	95.80	n/a
S96T004318	A	Lead - ICP-Acid Digest	ug/g	91.60	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	90.00	48.00	n/a
S96T004318	A	Sulfur - ICP-Acid Digest	ug/g	96.60	1.10e-01	2.00e+03	2.15e+03	2.08e+03	7.23	125.2	48.00	n/a
S96T004318	A	Antimony - ICP-Acid Digest	ug/g	94.00	<6.00e-02	< 28.70	<2.70e1	n/a	n/a	91.40	28.80	n/a
S96T004318	A	Selenium - ICP-Acid Digest	ug/g	90.20	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	85.80	48.00	n/a
S96T004318	A	Silicon - ICP-Acid Digest	ug/g	118.4	3.48e-01	< 61.70	170.0	115.8	93.5	93.80	23.90	n/a
S96T004318	A	Samarium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 47.90	<4.51e1	n/a	n/a	89.00	48.00	n/a
S96T004318	A	Strontium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	< 4.790	<4.51e0	n/a	n/a	88.60	4.800	n/a
S96T004318	A	Titanium - ICP-Acid Digest	ug/g	87.80	<1.00e-02	< 4.790	<4.51e0	n/a	n/a	83.20	4.800	n/a
S96T004318	A	Thallium - ICP-Acid Digest	ug/g	89.60	<2.00e-01	< 95.80	<9.02e1	n/a	n/a	84.60	95.80	n/a
S96T004318	A	Uranium - ICP-Acid Digest	ug/g	92.80	<5.00e-01	<2.40e+02	<2.25e2	n/a	n/a	89.50	240.0	n/a
S96T004318	A	Vanadium - ICP-Acid Digest	ug/g	93.60	<5.00e-02	< 24.00	<2.25e1	n/a	n/a	87.40	23.90	n/a
S96T004318	A	Zinc - ICP-Acid Digest	ug/g	91.80	1.03e-01	< 43.20	40.10	41.65	7.44	87.40	4.800	n/a
S96T004319	A	Zirconium - ICP-Acid Digest	ug/g	96.20	<1.00e-02	< 4.790	<4.51e0	n/a	n/a	91.60	4.800	n/a
S96T004319	A	Acetate by IC-Dionex 4000/4500	ug/g	99.19	<2.00e-02	< 82.06	<8.34e1	n/a	n/a	82.26	82.07	n/a
S96T004319	A	Formate by IC-Dionex 4000/4500	ug/g	101.7	<2.00e-02	6.64e+02	715.0	689.8	7.40	63.0	85.16	n/a
S96T004319	A	Fluoride by IC-Dionex 4000/4500	ug/g	107.6	<1.30e-02	1.09e+03	1.09e+03	1.09e+03	0.00	119.8	53.34	n/a
S96T004319	A	Chloride by IC-Dionex 4000/4500	ug/g	107.8	<1.30e-02	4.54e+03	5.49e+03	5.01e+03	18.9	126.6	69.76	n/a
S96T004319	A	Nitrite by IC-Dionex 4000/4500	ug/g	108.3	<1.07e-01	5.64e+04	6.82e+04	6.23e+04	18.9	142.7	439.1	n/a
S96T004319	A	Bromide by Ion Chromatograph	ug/g	105.3	<1.26e-01	5.61e+02	576.0	568.4	2.64	93.89	517.0	n/a
S96T004319	A	Nitrate by IC-Dionex 4000/4500	ug/g	106.1	<1.19e-01	9.44e+04	9.93e+04	9.69e+04	5.06	168.4	573.9	n/a
S96T004319	A	Phosphate by IC-Dionex 4000/4500	ug/g	105.0	<1.19e-01	7.39e+02	573.0	656.1	25.3	101.3	487.8	n/a
S96T004319	A	Sulfate by IC-Dionex 4000/4500	ug/g	106.7	<1.36e-01	5.13e+03	5.28e+03	5.20e+03	2.88	103.6	557.6	n/a
S96T004319	A	Oxalate by IC-Dionex 4000/4500	ug/g	106.2	<1.05e-01	7.64e+03	8.02e+03	7.83e+03	4.85	107.0	430.9	n/a

CORE NUMBER: 153
SEGMENT #: 17

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961004428		Bulk Density of Sample	g/mL	n/a	n/a	1.520	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961004435		DSC Exotherm using Mettler	Joules/g	93.85	n/a	48.70	94.10	71.40	63.6	n/a	n/a	n/a
S961004435		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	83.20	161.0	122.1	63.7	n/a	n/a	n/a
S961004435		% Water by TGA using Mettler	%	98.28	n/a	44.70	38.20	41.45	15.7	n/a	n/a	n/a
S961004441		F Silver - ICP-Fusion	ug/g	99.20	<1.00e-02	<1.96e+02	<1.96e2	2.44e+04	1.23	89.30	196.0	n/a
S961004441		F Aluminum - ICP-Fusion	ug/g	98.40	<5.00e-02	2.42e+04	2.45e+04	n/a	n/a	89.30	980.0	n/a
S961004441		F Arsenic - ICP-Fusion	ug/g	103.0	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	101.0	1.96e+03	n/a
S961004441		F Boron - ICP-Fusion	ug/g	99.60	<5.00e-02	<9.82e+02	<9.82e2	n/a	n/a	95.50	980.0	n/a
S961004441		F Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<9.80e+02	<9.82e2	n/a	n/a	94.40	980.0	n/a
S961004441		F Beryllium - ICP-Fusion	ug/g	102.6	<5.00e-03	<98.00	<9.82e1	n/a	n/a	97.60	98.00	n/a
S961004441		F Bismuth - ICP-Fusion	ug/g	99.80	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	96.80	1.96e+03	n/a
S961004441		F Calcium - ICP-Fusion	ug/g	100.4	<5.00e-03	<98.00	<9.82e1	n/a	n/a	99.20	1.96e+03	n/a
S961004441		F Cadmium - ICP-Fusion	ug/g	99.80	<5.00e-03	<98.00	<9.82e1	n/a	n/a	97.30	98.00	n/a
S961004441		F Cerium - ICP-Fusion	ug/g	99.80	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	96.30	1.96e+03	n/a
S961004441		F Cobalt - ICP-Fusion	ug/g	100.2	<2.00e-02	<3.92e+02	<3.92e2	n/a	n/a	97.10	392.0	n/a
S961004441		F Chromium - ICP-Fusion	ug/g	100.4	<1.00e-02	<3.71e+02	<3.41e+03	3.56e+03	8.43	97.50	196.0	n/a
S961004441		F Copper - ICP-Fusion	ug/g	103.8	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	98.80	196.0	n/a
S961004441		F Iron - ICP-Fusion	ug/g	100.0	<5.00e-02	<9.80e+02	<9.82e2	n/a	n/a	97.20	980.0	n/a
S961004441		F Potassium - ICP-Fusion	ug/g	99.80	<5.00e-01	5.37e+06	n/a	n/a	n/a	n/a	9.80e+03	n/a
S961004441		F Lanthanum - ICP-Fusion	ug/g	101.4	<5.00e-02	<1.96e+02	<1.96e2	n/a	n/a	97.50	980.0	n/a
S961004441		F Lithium - ICP-Fusion	ug/g	102.8	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	96.30	196.0	n/a
S961004441		F Magnesium - ICP-Fusion	ug/g	94.20	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	92.10	1.96e+03	n/a
S961004441		F Manganese - ICP-Fusion	ug/g	97.80	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	93.20	196.0	n/a
S961004441		F Molybdenum - ICP-Fusion	ug/g	100.4	<5.00e-02	<9.80e+02	<9.82e2	n/a	n/a	98.50	980.0	n/a
S961004441		F Sodium - ICP-Fusion	ug/g	98.60	1.600	2.23e+05	2.25e+05	2.24e+05	0.89	51.10	1.96e+03	n/a
S961004441		F Neodymium - ICP-Fusion	ug/g	101.6	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	97.50	1.96e+03	n/a
S961004441		F Nickel - ICP-Fusion	ug/g	99.20	3.80e-02	4.13e+02	<3.93e2	n/a	n/a	95.30	392.0	n/a
S961004441		F Phosphorus - ICP-Fusion	ug/g	105.0	<2.00e-01	<3.92e+03	<3.93e3	n/a	n/a	105.0	3.92e+03	n/a
S961004441		F Lead - ICP-Fusion	ug/g	99.00	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	98.00	1.96e+03	n/a
S961004441		F Sulfur - ICP-Fusion	ug/g	100.0	<1.00e-01	3.04e+03	2.64e+03	2.84e+03	14.1	96.60	1.96e+03	n/a
S961004441		F Antimony - ICP-Fusion	ug/g	95.60	<6.00e-02	<1.18e+03	<1.18e3	n/a	n/a	93.50	1.8e+03	n/a
S961004441		F Selenium - ICP-Fusion	ug/g	96.60	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	94.20	1.96e+03	n/a
S961004441		F Silicon - ICP-Fusion	ug/g	100.8	<5.00e-02	<9.80e+02	<9.82e2	n/a	n/a	104.0	980.0	n/a
S961004441		F Samarium - ICP-Fusion	ug/g	98.60	<1.00e-01	<1.96e+03	<1.96e3	n/a	n/a	94.30	1.96e+03	n/a
S961004441		F Strontium - ICP-Fusion	ug/g	98.60	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	94.50	196.0	n/a
S961004441		F Titanium - ICP-Fusion	ug/g	98.00	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	94.80	196.0	n/a
S961004441		F Thallium - ICP-Fusion	ug/g	96.20	<2.00e-01	<3.92e+03	<3.93e3	n/a	n/a	93.60	3.92e+03	n/a
S961004441		F Uranium - ICP-Fusion	ug/g	96.30	<5.00e-01	<9.80e+03	<9.82e3	n/a	n/a	89.50	9.80e+03	n/a
S961004441		F Vanadium - ICP-Fusion	ug/g	99.80	<5.00e-02	<9.80e+02	<9.82e2	n/a	n/a	96.60	980.0	n/a
S961004441		F Zinc - ICP-Fusion	ug/g	100.8	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	95.30	196.0	n/a
S961004441		F Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<1.96e+02	<1.96e2	n/a	n/a	97.20	196.0	n/a
S961004444		A Silver - ICP-Acid Digest	ug/g	44.70	<1.00e-02	12.60	13.20	12.90	4.65	n/a	4.830	n/a
S961004444		A Aluminum - ICP-Acid Digest	ug/g	98.80	1.22e-01	2.24e+04	2.29e+04	2.26e+04	2.21	n/a	24.10	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T004444	A	Arsenic - ICP-Acid Digest	ug/g	97.80	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Boron - ICP-Acid Digest	ug/g	106.8	4.15e-01	1.03e+02	147.0	125.0	35.2	N/A	24.10	N/A
S96T004444	A	Barium - ICP-Acid Digest	ug/g	96.00	<5.00e-02	< 24.20	<2.37e1	N/A	N/A	N/A	24.10	N/A
S96T004444	A	Beryllium - ICP-Acid Digest	ug/g	99.40	<5.00e-03	< 2.40	<4.75e1	N/A	N/A	N/A	2.420	N/A
S96T004444	A	Bismuth - ICP-Acid Digest	ug/g	94.80	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Calcium - ICP-Acid Digest	ug/g	180.8	4.180	6.24e+02	624.0	624.0	0.00	N/A	48.20	N/A
S96T004444	A	Cadmium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	< 10.80	12.50	11.65	14.6	N/A	2.420	N/A
S96T004444	A	Cerium - ICP-Acid Digest	ug/g	98.00	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Cobalt - ICP-Acid Digest	ug/g	95.60	<2.00e-02	< 9.660	<9.50e0	N/A	N/A	N/A	9.680	N/A
S96T004444	A	Chromium - ICP-Acid Digest	ug/g	94.40	<1.00e-02	2.69e+03	2.65e+03	2.67e+03	1.50	N/A	4.830	N/A
S96T004444	A	Copper - ICP-Acid Digest	ug/g	94.40	<1.00e-02	7.370	11.40	9.385	42.9	N/A	4.830	N/A
S96T004444	A	Potassium - ICP-Acid Digest	ug/g	94.80	<5.00e-02	60.20	62.30	61.25	3.43	N/A	24.10	N/A
S96T004444	A	Lanthanum - ICP-Acid Digest	ug/g	97.80	<5.00e-01	3.65e+03	3.63e+03	3.64e+03	0.55	N/A	24.20	N/A
S96T004444	A	Lithium - ICP-Acid Digest	ug/g	98.20	<5.00e-02	< 24.20	<2.37e1	N/A	N/A	N/A	24.10	N/A
S96T004444	A	Magnesium - ICP-Acid Digest	ug/g	90.80	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Manganese - ICP-Acid Digest	ug/g	92.20	<1.00e-02	7.520	8.620	8.070	13.6	N/A	4.830	N/A
S96T004444	A	Molybdenum - ICP-Acid Digest	ug/g	93.00	<5.00e-02	50.90	51.90	51.40	1.95	N/A	24.10	N/A
S96T004444	A	Sodium - ICP-Acid Digest	ug/g	134.4	8.91e-01	1.73e+05	1.73e+05	1.72e+05	0.00	N/A	48.20	N/A
S96T004444	A	Neodymium - ICP-Acid Digest	ug/g	97.00	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Nickel - ICP-Acid Digest	ug/g	98.80	<2.00e-02	46.60	46.60	46.60	0.00	N/A	9.680	N/A
S96T004444	A	Phosphorus - ICP-Acid Digest	ug/g	98.80	<2.00e-01	5.77e+02	518.0	547.5	10.8	N/A	96.80	N/A
S96T004444	A	Lead - ICP-Acid Digest	ug/g	91.60	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Sulfur - ICP-Acid Digest	ug/g	96.60	1.10e-01	2.23e+03	2.22e+03	2.22e+03	0.45	N/A	48.20	N/A
S96T004444	A	Antimony - ICP-Acid Digest	ug/g	90.20	<5.00e-02	< 29.00	<2.85e1	N/A	N/A	N/A	29.00	N/A
S96T004444	A	Selenium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Silicon - ICP-Acid Digest	ug/g	118.4	3.48e-01	1.59e+02	153.0	156.0	3.85	N/A	24.10	N/A
S96T004444	A	Samarium - ICP-Acid Digest	ug/g	95.00	<1.00e-01	< 48.30	<4.75e1	N/A	N/A	N/A	48.20	N/A
S96T004444	A	Strontium - ICP-Acid Digest	ug/g	95.20	<1.00e-02	< 4.830	<4.75e0	N/A	N/A	N/A	4.830	N/A
S96T004444	A	Titanium - ICP-Acid Digest	ug/g	87.80	<1.00e-01	< 4.830	<4.75e0	N/A	N/A	N/A	4.830	N/A
S96T004444	A	Thallium - ICP-Acid Digest	ug/g	89.60	<2.00e-01	< 96.60	<9.50e1	N/A	N/A	N/A	96.80	N/A
S96T004444	A	Uranium - ICP-Acid Digest	ug/g	92.80	<5.00e-01	<2.42e+02	<2.37e2	N/A	N/A	N/A	24.20	N/A
S96T004444	A	Vanadium - ICP-Acid Digest	ug/g	93.60	<5.00e-02	< 24.20	<2.37e1	N/A	N/A	N/A	24.10	N/A
S96T004444	A	Zinc - ICP-Acid Digest	ug/g	91.80	1.03e-01	42.30	41.80	42.05	1.19	N/A	4.830	N/A
S96T004444	A	Zirconium - ICP-Acid Digest	ug/g	96.20	<1.00e-02	5.580	6.530	6.055	15.7	N/A	4.830	N/A
S96T004447	A	Acetate by IC-BIOMEX 4000/4500	ug/g	99.19	<2.00e-02	3.89e+02	394.0	391.6	1.28	N/A	80.62	N/A
S96T004447	A	Formate by IC-BIOMEX 4000/4500	ug/g	101.7	<2.00e-02	8.02e+02	826.0	813.8	2.95	N/A	84.65	N/A
S96T004447	A	Fluoride-IC-BIOMEX 4000/4500	ug/g	107.6	<1.00e-02	1.21e+03	1.26e+03	1.28e+03	11.7	N/A	52.40	N/A
S96T004447	A	Chloride-IC-BIOMEX 4000/4500	ug/g	107.8	<1.00e-02	4.92e+03	6.09e+03	5.51e+03	21.3	N/A	68.52	N/A
S96T004447	A	Nitrite-IC - BIOMEX 4000/4500	ug/g	108.3	<1.07e-01	5.89e+04	7.27e+04	6.58e+04	21.0	N/A	431.3	N/A
S96T004447	A	Bromide by Ion Chromatograph	ug/g	105.3	<1.26e-01	5.16e+02	537.0	526.4	3.99	N/A	507.8	N/A
S96T004447	A	Nitrate by IC-BIOMEX 4000/4500	ug/g	106.1	<1.00e-01	7.70e+02	1.00e+03	8.85e+02	41.9	N/A	563.9	N/A
S96T004447	A	Phosphate-IC-BIOMEX 4000/4500	ug/g	105.0	<1.19e-01	9.87e+02	1.51e+03	1.25e+03	61.9	N/A	479.2	N/A
S96T004447	A	Sulfate by IC-BIOMEX 4000/4500	ug/g	106.7	<1.36e-01	5.92e+03	6.46e+03	6.19e+03	8.72	N/A	547.7	N/A
S96T004447	A	Oxalate by IC-BIOMEX 4000/4500	ug/g	106.2	<1.05e-01	9.76e+03	1.00e+04	9.88e+03	2.43	N/A	423.1	N/A

L Lower Half of Segment: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961004427		Bulk Density of Sample	g/mL	n/a	n/a	1.530	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S961004427		DSC Exotherm using Mettler	Joules/g	103.7	n/a	19.60	19.60	17.70	21.5	n/a	n/a	n/a	n/a
S961004436		DSC Exotherm Dry, Calculated	Joules/g Dry	n/a	n/a	28.00	34.70	31.35	21.4	n/a	n/a	n/a	n/a
S961004436		% Water by TGA using Mettler	%	99.19	n/a	44.54	42.57	43.55	4.52	n/a	n/a	n/a	n/a
S961004442		F Silver - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.01e-02	<2.00e2	2.44e+04	n/a	89.60	201.0	n/a	n/a
S961004442		F Aluminum - ICP-Fusion	ug/g	98.40	<5.00e-02	<2.46e+04	<2.43e+04	n/a	n/a	92.50	1.01e+03	n/a	n/a
S961004442		F Arsenic - ICP-Fusion	ug/g	103.0	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	102.0	1.01e+03	n/a	n/a
S961004442		F Boron - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	96.50	1.01e+03	n/a	n/a
S961004442		F Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	95.10	1.01e+03	n/a	n/a
S961004442		F Beryllium - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.01e+02	<2.00e2	n/a	n/a	98.20	101.0	n/a	n/a
S961004442		F Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	97.40	2.01e+03	n/a	n/a
S961004442		F Cadmium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	99.50	2.01e+03	n/a	n/a
S961004442		F Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	97.90	101.0	n/a	n/a
S961004442		F Cerium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	97.40	2.01e+03	n/a	n/a
S961004442		F Cobalt - ICP-Fusion	ug/g	100.2	<2.00e-02	<4.03e+02	<4.00e2	n/a	n/a	97.60	403.0	n/a	n/a
S961004442		F Chromium - ICP-Fusion	ug/g	100.4	<1.00e-02	<2.01e+02	<2.00e2	4.94e+03	5.86	98.20	201.0	n/a	n/a
S961004442		F Copper - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	97.60	1.01e+03	n/a	n/a
S961004442		F Iron - ICP-Fusion	ug/g	99.80	<5.00e-01	5.40e+06	n/a	n/a	n/a	98.10	1.01e+04	n/a	n/a
S961004442		F Potassium - ICP-Fusion	ug/g	101.4	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	96.10	1.01e+03	n/a	n/a
S961004442		F Lanthanum - ICP-Fusion	ug/g	102.8	<1.00e-02	<2.01e+02	<2.00e2	n/a	n/a	92.70	2.01e+03	n/a	n/a
S961004442		F Lithium - ICP-Fusion	ug/g	94.20	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	93.50	201.0	n/a	n/a
S961004442		F Magnesium - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.01e+02	<2.00e2	n/a	n/a	98.60	1.01e+03	n/a	n/a
S961004442		F Manganese - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	97.90	2.01e+03	n/a	n/a
S961004442		F Molybdenum - ICP-Fusion	ug/g	98.60	1.600	2.15e+05	2.20e+05	2.18e+05	2.30	75.90	2.01e+03	n/a	n/a
S961004442		F Sodium - ICP-Fusion	ug/g	101.6	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	97.80	2.01e+03	n/a	n/a
S961004442		F Neodymium - ICP-Fusion	ug/g	99.20	3.80e-02	4.59e+02	<4.00e2	n/a	n/a	96.10	403.0	n/a	n/a
S961004442		F Nickel - ICP-Fusion	ug/g	105.0	<2.00e-01	<4.03e+03	<4.00e3	n/a	n/a	96.20	2.01e+03	n/a	n/a
S961004442		F Phosphorus - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.01e+03	<2.00e3	n/a	n/a	97.40	2.01e+03	n/a	n/a
S961004442		F Lead - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.01e+03	<2.00e3	2.66e+03	0.38	97.40	2.01e+03	n/a	n/a
S961004442		F Sulfur - ICP-Fusion	ug/g	95.60	<6.00e-02	<1.21e+03	<1.20e3	n/a	n/a	94.80	1.21e+03	n/a	n/a
S961004442		F Antimony - ICP-Fusion	ug/g	96.60	<1.00e-01	<1.83e+03	<1.80e3	n/a	n/a	95.40	2.01e+03	n/a	n/a
S961004442		F Selenium - ICP-Fusion	ug/g	100.0	<5.00e-02	<2.01e+02	<2.00e2	n/a	n/a	99.90	1.01e+03	n/a	n/a
S961004442		F Samarium - ICP-Fusion	ug/g	98.60	<1.00e-01	<1.83e+03	<1.80e3	n/a	n/a	94.80	2.01e+03	n/a	n/a
S961004442		F Strontium - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.01e+02	<2.00e2	n/a	n/a	95.80	201.0	n/a	n/a
S961004442		F Titanium - ICP-Fusion	ug/g	98.20	<2.00e-01	<4.03e+03	<4.00e3	n/a	n/a	95.80	201.0	n/a	n/a
S961004442		F Vanadium - ICP-Fusion	ug/g	96.30	<5.00e-02	<1.01e+04	<1.00e4	n/a	n/a	90.00	1.01e+04	n/a	n/a
S961004442		F Uranium - ICP-Fusion	ug/g	99.80	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	97.30	1.01e+03	n/a	n/a
S961004442		F Zinc - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.01e+02	<2.00e2	n/a	n/a	99.50	201.0	n/a	n/a
S961004442		F Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.01e+02	<2.00e2	n/a	n/a	95.80	201.0	n/a	n/a
S961004442		F Alpha of Digested Solid	ug/g	103.1	<1.72e-03	7.15e-03	6.95e-03	7.04e-03	3.12	82.73	3.00e-03	5.19E+01	n/a
S961004445		A Silver - ICP-Acid Digest	ug/g	44.00	<9.00e-02	2.16e+04	14.50	2.15e+04	0.93	n/a	3.720	n/a	n/a
S961004445		A Aluminum - ICP-Acid Digest	ug/g	92.20	<1.00e-01	<3.71e-01	<3.66e-1	n/a	n/a	n/a	18.60	n/a	n/a
S961004445		A Arsenic - ICP-Acid Digest	ug/g	92.80	<1.00e-01	<3.71e-01	<3.66e-1	n/a	n/a	n/a	18.60	n/a	n/a
S961004445		A Boron - ICP-Acid Digest	ug/g	99.40	<4.89e-01	<115.0	<115.0	95.35	4.12	n/a	18.60	n/a	n/a
S961004445		A Barium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	<18.60	<18.60	n/a	n/a	n/a	1.860	n/a	n/a
S961004445		A Beryllium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	<1.860	<1.860	n/a	n/a	n/a	1.860	n/a	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec	Det Limit	Count Err%
S961004445	A	Bismuth - ICP-Acid Digest	ug/g	102.60	<1.00e-01	< 37.10	<3.66e1	n/a	n/a	n/a	37.20	n/a
S961004445	A	Calcium - ICP-Acid Digest	ug/g	94.61	2.27e-01	8.20e+02	611.0	715.5	29.2	n/a	37.20	n/a
S961004445	A	Cadmium - ICP-Acid Digest	ug/g	89.40	<5.00e-03	38.40	21.60	20.00	16.0	n/a	1.860	n/a
S961004445	A	Cerium - ICP-Acid Digest	ug/g	95.60	<1.00e-01	< 37.10	<3.66e1	n/a	n/a	n/a	37.20	n/a
S961004445	A	Cobalt - ICP-Acid Digest	ug/g	91.80	<2.00e-02	< 7.20	<7.32e0	n/a	n/a	n/a	7.420	n/a
S961004445	A	Chromium - ICP-Acid Digest	ug/g	90.40	<1.00e-02	4.22e+03	4.37e+03	4.30e+03	3.02	n/a	3.720	n/a
S961004445	A	Copper - ICP-Acid Digest	ug/g	89.80	<1.00e-02	8.150	13.70	10.93	5.08	n/a	3.720	n/a
S961004445	A	Iron - ICP-Acid Digest	ug/g	93.80	<1.20e-01	91.60	86.50	89.25	5.27	n/a	18.60	n/a
S961004445	A	Potassium - ICP-Acid Digest	ug/g	91.00	<5.00e-02	3.39e+01	3.32e+03	3.36e+03	2.09	n/a	186.0	n/a
S961004445	A	Lanthanum - ICP-Acid Digest	ug/g	93.60	<5.00e-02	18.60	<1.85e1	13.20	11.36	n/a	18.60	n/a
S961004445	A	Lithium - ICP-Acid Digest	ug/g	93.40	<1.00e-01	9.550	13.20	11.75	24.1	n/a	37.20	n/a
S961004445	A	Magnesium - ICP-Acid Digest	ug/g	86.20	<1.00e-01	37.10	<3.66e1	15.20	2.33	n/a	18.60	n/a
S961004445	A	Manganese - ICP-Acid Digest	ug/g	90.60	<5.00e-02	50.90	52.10	51.50	2.33	n/a	37.20	n/a
S961004445	A	Molybdenum - ICP-Acid Digest	ug/g	109.8	8.47e-01	1.67e+05	1.66e+05	1.66e+05	0.60	n/a	37.20	n/a
S961004445	A	Sodium - ICP-Acid Digest	ug/g	92.60	<1.00e-01	< 37.10	<3.66e1	n/a	n/a	n/a	37.20	n/a
S961004445	A	Neodymium - ICP-Acid Digest	ug/g	92.20	<2.00e-02	78.40	83.00	80.70	3.70	n/a	7.420	n/a
S961004445	A	Nickel - ICP-Acid Digest	ug/g	93.00	<2.00e-01	7.99e+02	785.0	792.0	1.77	n/a	74.50	n/a
S961004445	A	Phosphorus - ICP-Acid Digest	ug/g	87.80	<1.00e-01	45.40	56.10	50.75	2.11	n/a	37.20	n/a
S961004445	A	Lead - ICP-Acid Digest	ug/g	91.60	<1.00e-01	2.31e+03	2.36e+03	2.34e+03	2.14	n/a	37.20	n/a
S961004445	A	Sulfur - ICP-Acid Digest	ug/g	89.60	<6.00e-02	< 22.30	<2.20e1	n/a	n/a	n/a	22.50	n/a
S961004445	A	Antimony - ICP-Acid Digest	ug/g	85.20	<1.00e-01	< 37.10	<3.66e1	n/a	n/a	n/a	37.20	n/a
S961004445	A	Selenium - ICP-Acid Digest	ug/g	118.2	2.57e-01	1.90e+02	176.0	183.0	7.65	n/a	18.60	n/a
S961004445	A	Silicon - ICP-Acid Digest	ug/g	90.40	<1.00e-01	< 37.10	<3.66e1	n/a	n/a	n/a	37.20	n/a
S961004445	A	Samarium - ICP-Acid Digest	ug/g	90.80	<1.00e-02	< 3.70	3.750	n/a	n/a	n/a	37.20	n/a
S961004445	A	Strontium - ICP-Acid Digest	ug/g	83.60	<1.00e-02	< 3.70	<3.66e0	n/a	n/a	n/a	37.20	n/a
S961004445	A	Titanium - ICP-Acid Digest	ug/g	84.80	<2.00e-01	74.20	<7.32e1	n/a	n/a	n/a	18.60	n/a
S961004445	A	Thallium - ICP-Acid Digest	ug/g	88.60	<5.00e-01	<1.86e+02	<1.83e2	n/a	n/a	n/a	186.0	n/a
S961004445	A	Uranium - ICP-Acid Digest	ug/g	89.80	<5.00e-02	< 18.60	<1.83e1	n/a	n/a	n/a	18.60	n/a
S961004445	A	Vanadium - ICP-Acid Digest	ug/g	89.40	<1.00e-01	30.50	42.10	36.30	32.0	n/a	37.20	n/a
S961004445	A	Zinc - ICP-Acid Digest	ug/g	91.40	<1.00e-02	8.090	12.10	10.09	39.7	n/a	37.20	n/a
S961004445	A	Zirconium - ICP-Acid Digest	ug/g	107.3	<2.00e-02	7.66e+02	805.0	785.3	4.96	80.48	41.84	n/a
S961004448	W	Acetate by IC-DIONEX 4000/4500	ug/g	101.7	<2.00e-02	1.02e+03	1.13e+03	1.07e+03	10.2	116.7	43.95	n/a
S961003325	W	Formate by IC-DIONEX 4000/4500	ug/g	94.36	<2.00e-02	5.37e+02	501.0	519.0	6.94	125.0	77.45	n/a
S961003325	W	Acetate by IC-DIONEX 4000/4500	ug/g	98.33	<2.00e-02	1.19e+03	1.06e+03	1.12e+03	11.6	99.17	81.31	n/a
S961003325	W	Formate by IC-DIONEX 4000/4500	ug/g	103.2	<1.00e-02	1.22e+02	1.26e+03	1.24e+03	3.23	143.6	46.45	n/a
S961003325	W	Fluoride-IC-DIONEX 4000/4500	ug/g	103.5	<1.70e-02	6.72e+02	6.26e+03	6.49e+03	7.09	99.49	65.81	n/a
S961003325	W	Chloride-IC-DIONEX 4000/4500	ug/g	97.79	<1.00e-01	7.86e+04	7.38e+04	7.62e+04	6.30	100.7	418.1	n/a
S961003325	W	Nitrite-IC - DIONEX 4000/4500	ug/g	98.30	<1.25e-01	<4.84e+01	<4.78e2	n/a	n/a	93.04	484.1	n/a
S961003325	W	Bromide by Ion Chromatograph	ug/g	99.33	<1.39e-01	1.05e+05	9.77e+04	1.01e+05	7.20	100.0	538.2	n/a
S961003325	W	Nitrate by IC-DIONEX 4000/4500	ug/g	99.45	<1.20e-01	2.00e+03	2.68e+03	2.34e+03	29.1	97.43	464.5	n/a
S961003325	W	Phosphate by IC-DIONEX 4000/4500	ug/g	99.99	<1.38e-01	6.64e+03	6.39e+03	6.51e+03	3.84	96.35	534.2	n/a
S961003325	W	Sulfate by IC-DIONEX 4000/4500	ug/g	99.62	<1.09e-01	1.18e+04	1.22e+04	1.20e+04	3.33	99.81	406.6	n/a
S961003325	W	Oxalate by IC-DIONEX 4000/4500	ug/g									

Table 1. Data Summary Table
AN-105

CORE NUMBER: 153
SEGMENT #: 19

SEGMENT PORTION: L Lower Half of Segment

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961004429		Bulk Density of Sample	g/mL			1.560	n/a	n/a	n/a	n/a	5.0e-01	n/a
S961004437		DSC Exotherm using Mettler	Joules/g	103.7	n/a	22.90	44.40	33.65	63.9	n/a	n/a	n/a
S961004437		DSC Exotherm DRY Calculated	Joules/g DRY			81.30	81.30	61.60	64.0	n/a	n/a	n/a
S961004437		% Water by TGA using Mettler	%	98.19	n/a	46.16	44.59	45.38	3.46	n/a	n/a	n/a
S961004443	F	Silver - ICP-Fusion	ug/g	99.20	<1.00e-02	2.32e+02	<2.00e2	87.00	82.00	n/a	202.0	n/a
S961004443	F	Aluminum - ICP-Fusion	ug/g	98.40	<5.00e-02	2.43e+04	2.47e+04	2.42e+04	4.55	92.90	1.01e+03	n/a
S961004443	F	Arsenic - ICP-Fusion	ug/g	103.0	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	101.40	1.01e+03	n/a
S961004443	F	Boron - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	95.40	1.01e+03	n/a
S961004443	F	Barium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	97.90	101.0	n/a
S961004443	F	Beryllium - ICP-Fusion	ug/g	102.6	<5.00e-03	<1.01e+02	<1.00e2	n/a	n/a	95.70	2.02e+03	n/a
S961004443	F	Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	98.60	2.02e+03	n/a
S961004443	F	Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	96.10	101.0	n/a
S961004443	F	Cadmium - ICP-Fusion	ug/g	99.80	<5.00e-03	<1.01e+02	<1.00e2	n/a	n/a	96.10	2.02e+03	n/a
S961004443	F	Cerium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	97.00	202.0	n/a
S961004443	F	Cobalt - ICP-Fusion	ug/g	100.2	<2.00e-02	<4.04e+02	<4.00e2	3.61e+03	15.5	97.00	202.0	n/a
S961004443	F	Chromium - ICP-Fusion	ug/g	103.8	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	98.70	202.0	n/a
S961004443	F	Copper - ICP-Fusion	ug/g	100.0	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	96.80	1.01e+03	n/a
S961004443	F	Iron - ICP-Fusion	ug/g	99.80	<5.00e-01	5.57e+06	n/a	n/a	n/a	97.50	1.01e+03	n/a
S961004443	F	Potassium - ICP-Fusion	ug/g	101.4	<5.00e-02	<2.02e+02	<2.00e2	n/a	n/a	96.90	202.0	n/a
S961004443	F	Lanthanum - ICP-Fusion	ug/g	102.8	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	91.90	2.02e+03	n/a
S961004443	F	Magnesium - ICP-Fusion	ug/g	94.20	<1.00e-01	<2.02e+02	<2.00e2	n/a	n/a	92.50	202.0	n/a
S961004443	F	Manganese - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	98.90	2.02e+03	n/a
S961004443	F	Molybdenum - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.01e+02	<1.00e3	n/a	n/a	98.50	1.01e+03	n/a
S961004443	F	Sodium - ICP-Fusion	ug/g	98.60	1.600	2.19e+05	2.21e+05	2.20e+05	0.91	81.30	2.02e+03	n/a
S961004443	F	Neodymium - ICP-Fusion	ug/g	101.6	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	97.30	2.02e+03	n/a
S961004443	F	Nickel - ICP-Fusion	ug/g	99.20	3.80e-02	<4.04e+02	<4.00e2	n/a	n/a	97.10	404.0	n/a
S961004443	F	Phosphorus - ICP-Fusion	ug/g	105.0	<2.00e-01	<4.04e+03	<4.00e3	n/a	n/a	105.0	4.04e+03	n/a
S961004443	F	Lead - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.02e+03	<2.00e3	2.96e+03	16.5	96.40	2.02e+03	n/a
S961004443	F	Sulfur - ICP-Fusion	ug/g	95.60	<5.00e-02	<1.21e+03	<1.20e3	n/a	n/a	93.40	1.21e+03	n/a
S961004443	F	Antimony - ICP-Fusion	ug/g	96.60	<1.00e-01	<2.02e+03	<2.00e3	n/a	n/a	91.80	2.02e+03	n/a
S961004443	F	Selenium - ICP-Fusion	ug/g	98.60	<1.00e-02	<2.02e+03	<2.00e3	n/a	n/a	103.0	1.01e+03	n/a
S961004443	F	Samarium - ICP-Fusion	ug/g	100.8	<5.00e-02	<2.02e+03	<2.00e3	n/a	n/a	93.90	2.02e+03	n/a
S961004443	F	Strontium - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.02e+02	<2.00e2	n/a	n/a	94.40	202.0	n/a
S961004443	F	Titanium - ICP-Fusion	ug/g	98.00	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	94.80	4.04e+03	n/a
S961004443	F	Thallium - ICP-Fusion	ug/g	96.20	<5.00e-01	<1.01e+04	<1.00e4	n/a	n/a	89.50	1.01e+04	n/a
S961004443	F	Uranium - ICP-Fusion	ug/g	99.80	<5.00e-02	<1.01e+03	<1.00e3	n/a	n/a	96.40	1.01e+03	n/a
S961004443	F	Vanadium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	98.80	202.0	n/a
S961004443	F	Zinc - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	95.10	202.0	n/a
S961004443	F	Zirconium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.02e+02	<2.00e2	n/a	n/a	98.00	3.00e-03	4.46e+01
S961004443	F	Alpha of Digested Solid	ug/g	103.1	<1.72e-03	1.13e-02	1.45e-02	43.6	n/a	n/a	n/a	n/a
S961004446	A	Silver - ICP-Acid Digest	ug/g	44.00	<1.00e-02	13.30	14.00	13.65	5.13	n/a	3.820	n/a

HNF-SD-WM-07-199, 199

WHC-SD-WM-DF-199 REV. 1

Sample #	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961004446	A	Aluminum - ICP-Acid Digest	ug/g	92.20	9.00e-02	2.09e+04	2.18e+04	2.10e+04	1.43	n/a	19.10	n/a
S961004446	A	Arsenic - ICP-Acid Digest	ug/g	92.80	<1.00e-01	<3.58e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Boron - ICP-Acid Digest	ug/g	99.40	4.80e-02	1.19e+02	108.0	113.5	9.69	n/a	19.10	n/a
S961004446	A	Barium - ICP-Acid Digest	ug/g	91.80	<5.00e-01	<1.90e-02	<1.90e-02	<1.90e-02	n/a	n/a	19.10	n/a
S961004446	A	Beryllium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	<1.91e-01	<1.91e-01	<1.91e-01	n/a	n/a	1.910	n/a
S961004446	A	Bismuth - ICP-Acid Digest	ug/g	93.60	<1.00e-01	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Calcium - ICP-Acid Digest	ug/g	142.8	2.270	1.47e+03	1.12e+03	1.30e+03	27.0	n/a	38.20	n/a
S961004446	A	Cadmium - ICP-Acid Digest	ug/g	89.40	<5.00e-03	13.30	13.70	13.50	2.96	n/a	1.910	n/a
S961004446	A	Cerium - ICP-Acid Digest	ug/g	93.60	<1.00e-01	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Cobalt - ICP-Acid Digest	ug/g	91.80	<2.00e-02	<7.65e-02	<7.65e-02	<7.65e-02	n/a	n/a	7.640	n/a
S961004446	A	Chromium - ICP-Acid Digest	ug/g	90.40	<1.00e-02	2.98e+03	3.16e+03	3.07e+03	5.86	n/a	3.820	n/a
S961004446	A	Copper - ICP-Acid Digest	ug/g	89.80	<1.00e-02	7.110	6.870	6.990	3.43	n/a	3.820	n/a
S961004446	A	Iron - ICP-Acid Digest	ug/g	93.80	1.20e-01	1.01e+02	70.90	85.95	35.0	n/a	19.10	n/a
S961004446	A	Potassium - ICP-Acid Digest	ug/g	91.00	<5.00e-01	3.20e+03	3.27e+03	3.24e+03	2.16	n/a	19.10	n/a
S961004446	A	Lanthanum - ICP-Acid Digest	ug/g	93.60	<5.00e-02	<1.91e-01	<1.91e-01	<1.91e-01	n/a	n/a	19.10	n/a
S961004446	A	Lithium - ICP-Acid Digest	ug/g	93.40	<1.00e-02	14.60	15.40	15.00	5.33	n/a	3.820	n/a
S961004446	A	Magnesium - ICP-Acid Digest	ug/g	86.20	<1.00e-01	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Manganese - ICP-Acid Digest	ug/g	88.40	<1.00e-02	10.50	10.40	10.45	0.96	n/a	3.820	n/a
S961004446	A	Molybdenum - ICP-Acid Digest	ug/g	90.60	<5.00e-02	49.10	47.40	48.25	3.52	n/a	19.10	n/a
S961004446	A	Sodium - ICP-Acid Digest	ug/g	109.8	8.41e-01	1.71e+05	1.79e+05	1.75e+05	4.57	n/a	38.20	n/a
S961004446	A	Neodymium - ICP-Acid Digest	ug/g	92.60	<1.00e-02	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Nickel - ICP-Acid Digest	ug/g	92.20	<2.00e-02	58.70	60.70	59.70	3.35	n/a	7.640	n/a
S961004446	A	Phosphorus - ICP-Acid Digest	ug/g	93.00	<2.00e-01	7.31e+02	2.02e+03	1.38e+03	93.7	n/a	38.20	n/a
S961004446	A	Lead - ICP-Acid Digest	ug/g	87.80	<1.00e-01	43.20	41.70	42.45	3.53	n/a	38.20	n/a
S961004446	A	Sulfur - ICP-Acid Digest	ug/g	91.60	<1.00e-01	2.47e+03	2.57e+03	2.52e+03	3.97	n/a	23.00	n/a
S961004446	A	Antimony - ICP-Acid Digest	ug/g	89.60	<6.00e-02	<22.90	<2.29e1	<2.29e1	n/a	n/a	38.20	n/a
S961004446	A	Selenium - ICP-Acid Digest	ug/g	85.20	<1.00e-01	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	19.10	n/a
S961004446	A	Silicon - ICP-Acid Digest	ug/g	118.2	2.57e-01	1.26e+02	164.0	145.0	26.2	n/a	19.10	n/a
S961004446	A	Samarium - ICP-Acid Digest	ug/g	90.40	<1.00e-02	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	38.20	n/a
S961004446	A	Strontium - ICP-Acid Digest	ug/g	90.80	<1.00e-02	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	3.820	n/a
S961004446	A	Titanium - ICP-Acid Digest	ug/g	83.60	<1.00e-02	<3.82e-02	<3.82e-02	<3.82e-02	n/a	n/a	3.820	n/a
S961004446	A	Thallium - ICP-Acid Digest	ug/g	84.80	<2.00e-01	<76.50	<7.65e1	<7.65e1	n/a	n/a	76.40	n/a
S961004446	A	Uranium - ICP-Acid Digest	ug/g	88.60	<5.00e-01	<1.91e+02	<1.91e+02	<1.91e+02	n/a	n/a	19.10	n/a
S961004446	A	Vanadium - ICP-Acid Digest	ug/g	89.80	<5.00e-02	<1.91e-02	<1.91e-02	<1.91e-02	n/a	n/a	19.10	n/a
S961004446	A	Zinc - ICP-Acid Digest	ug/g	91.40	<1.00e-01	66.80	43.90	55.35	41.4	n/a	3.820	n/a
S961004446	A	Zirconium - ICP-Acid Digest	ug/g	89.40	<1.00e-02	6.610	7.080	6.845	6.87	n/a	3.820	n/a
S961004449	W	Acetate by IC-DIONEX 4000/4500	ug/g	107.3	<2.10e-02	9.14e+02	897.0	905.7	1.88	n/a	42.53	n/a
S961005326	W	Formate by IC-DIONEX 4000/4500	ug/g	101.7	<2.10e-02	1.11e+03	1.09e+03	1.10e+03	1.82	n/a	44.66	n/a
S961005326	W	Acetate by IC-DIONEX 4000/4500	ug/g	94.36	<2.00e-02	5.36e+02	533.0	534.7	0.56	n/a	77.70	n/a
S961005326	W	Formate by IC-DIONEX 4000/4500	ug/g	98.33	<2.10e-02	1.14e+03	1.14e+03	1.14e+03	0.00	n/a	81.58	n/a
S961005326	W	Fluoride-IC-DIONEX 4000/4500	ug/g	103.2	<1.20e-02	1.09e+03	1.32e+03	1.21e+03	19.1	n/a	46.42	n/a
S961005326	W	Chloride-IC-DIONEX 4000/4500	ug/g	103.5	<1.70e-02	6.53e+03	6.17e+03	6.35e+03	5.67	n/a	66.05	n/a
S961005326	W	Nitrite-IC - Dionex 4000/4500	ug/g	97.79	<1.00e-01	7.52e+04	7.26e+04	7.39e+04	3.52	n/a	419.5	n/a
S961005326	W	Bromide by Ion Chromatograph	ug/g	98.30	<1.25e-01	<4.86e+02	<4.91e2	<4.91e2	n/a	n/a	485.5	n/a
S961005326	W	Nitrate by IC-DIONEX 4000/4500	ug/g	99.33	<1.39e-01	1.14e+05	1.27e+05	1.20e+05	10.8	n/a	589.2	n/a
S961005326	W	Phosphate-IC-DIONEX 4000/4500	ug/g	99.45	<1.20e-01	3.32e+03	5.50e+03	4.41e+03	49.4	n/a	446.2	n/a
S961005326	W	Sulfate by IC-DIONEX 4000/4500	ug/g	96.99	<1.38e-01	5.46e+03	5.91e+03	5.68e+03	7.92	n/a	536.1	n/a
S961005326	W	Oxalate by IC-DIONEX 4000/4500	ug/g	99.62	<1.05e-01	5.63e+03	6.64e+03	6.14e+03	16.5	n/a	407.8	n/a

Table 1. Data Summary Table
AN-105CORE NUMBER: 153
SEGMENT #: 21

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD, %	Spk Rec. %	Det Limit	Count Err%
S961003971		Bulk Density of Sample	g/mL	n/a	n/a	1.570	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961003981		DSC Exotherm using Mettler	Joules/g	102.6	n/a	21.00	33.20	19.60	14.3	n/a	n/a	n/a
S961003981		% Water by IGA using Mettler	Joules/g DRY	n/a	n/a	38.40	35.30	35.85	14.2	n/a	n/a	n/a
S961003981		% Water by IGA using Mettler	%	99.53	n/a	44.99	45.60	45.30	1.35	n/a	n/a	n/a
S961003984		F Silver - ICP-Fusion	ug/g	100.61	<1.00e-02	<2.01e-02	<2.01e-02	2.26e+04	3.10	85.30	201.0	n/a
S961003984		F Aluminum - ICP-Fusion	ug/g	98.80	<5.00e-02	2.22e-04	<2.01e-03	2.26e+04	3.10	103.0	1.00e+03	n/a
S961003984		F Arsenic - ICP-Fusion	ug/g	104.2	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	103.0	1.00e+03	n/a
S961003984		F Boron - ICP-Fusion	ug/g	99.80	<5.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	96.50	1.00e+03	n/a
S961003984		F Barium - ICP-Fusion	ug/g	102.8	<5.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	101.0	100.0	n/a
S961003984		F Beryllium - ICP-Fusion	ug/g	103.2	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	101.0	2.01e+03	n/a
S961003984		F Bismuth - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	101.0	100.0	n/a
S961003984		F Cadmium - ICP-Fusion	ug/g	101.6	<5.00e-03	<2.01e-02	<2.01e-02	n/a	n/a	98.60	2.01e+03	n/a
S961003984		F Cerium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	101.0	100.0	n/a
S961003984		F Cobalt - ICP-Fusion	ug/g	103.0	<2.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	101.0	201.0	n/a
S961003984		F Chromium - ICP-Fusion	ug/g	101.8	<1.00e-02	1.48e+02	1.43e+03	1.44e+03	3.44	100.0	201.0	n/a
S961003984		F Copper - ICP-Fusion	ug/g	101.6	<5.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	100.0	1.00e+03	n/a
S961003984		F Iron - ICP-Fusion	ug/g	93.00	<5.00e-01	5.91e+06	<2.01e-03	n/a	n/a	99.70	1.00e+03	n/a
S961003984		F Potassium - ICP-Fusion	ug/g	101.6	<5.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	99.60	201.0	n/a
S961003984		F Lanthanum - ICP-Fusion	ug/g	102.2	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	95.70	2.01e+03	n/a
S961003984		F Lithium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.01e-02	<2.01e-02	n/a	n/a	95.40	201.0	n/a
S961003984		F Magnesium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	102.0	1.00e+03	n/a
S961003984		F Manganese - ICP-Fusion	ug/g	103.4	<5.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	99.80	2.01e+03	n/a
S961003984		F Molybdenum - ICP-Fusion	ug/g	99.40	1.640	2.28e+05	2.24e+05	2.26e+05	1.77	92.80	2.01e+03	n/a
S961003984		F Sodium - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	101.0	403.0	n/a
S961003984		F Neodymium - ICP-Fusion	ug/g	100.8	7.80e-02	4.65e+02	4.09e+02	637.0	54.0	110.0	4.03e+03	n/a
S961003984		F Nickel - ICP-Fusion	ug/g	106.8	<2.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	99.30	2.01e+03	n/a
S961003984		F Lead - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	98.10	1.21e+03	n/a
S961003984		F Selenium - ICP-Fusion	ug/g	98.20	<6.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	97.20	2.01e+03	n/a
S961003984		F Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	101.0	1.00e+03	n/a
S961003984		F Strontium - ICP-Fusion	ug/g	96.80	<5.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	96.50	2.01e+03	n/a
S961003984		F Silicon - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	97.00	201.0	n/a
S961003984		F Samarium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	98.20	201.0	n/a
S961003984		F Strontium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	97.00	4.03e+03	n/a
S961003984		F Tantalum - ICP-Fusion	ug/g	97.40	<2.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	92.50	1.00e+04	n/a
S961003984		F Thallium - ICP-Fusion	ug/g	97.70	<5.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	103.0	201.0	n/a
S961003984		F Vanadium - ICP-Fusion	ug/g	102.2	<5.00e-02	<2.01e-03	<2.01e-03	n/a	n/a	98.40	201.0	n/a
S961003984		F Zinc - ICP-Fusion	ug/g	103.6	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	75.77	4.00e-03	5.0ME+01
S961003984		F Zirconium - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.01e-02	<2.01e-02	n/a	n/a	50.40	3.560	n/a
S961003984		F Alpha of Digested Solid	ug/g	102.3	<3.15e-03	8.49e-03	7.42e-03	13.40	13.55	50.40	3.560	n/a
S961003988	A	Silver - ICP-Acid Digest	ug/g	44.00	<1.00e-02	13.70	13.40	13.40	2.21	50.40	3.560	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Rec %	Det Limit	Count Err%
S961003988	A	Aluminum - ICP-Acid Digest	ug/g	92.20	9.00e-02	1.92e+04	1.90e+04	1.91e+04	1.05	143.4	17.80	n/a
S961003988	A	Arsenic - ICP-Acid Digest	ug/g	92.80	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	83.40	35.60	n/a
S961003988	A	Boron - ICP-Acid Digest	ug/g	99.40	4.89e-01	81.90	102.0	91.95	21.9	87.00	17.80	n/a
S961003988	A	Barium - ICP-Acid Digest	ug/g	91.80	<5.00e-02	< 17.80	<1.78e1	n/a	n/a	83.00	17.80	n/a
S961003988	A	Beryllium - ICP-Acid Digest	ug/g	93.60	<5.00e-03	< 1.78e1	<1.78e1	n/a	n/a	83.40	17.80	n/a
S961003988	A	Bismuth - ICP-Acid Digest	ug/g	90.60	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	82.60	35.60	n/a
S961003988	A	Calcium - ICP-Acid Digest	ug/g	142.6	2.270	6.32e+02	6.41e1	636.5	1.41	69.20	35.60	n/a
S961003988	A	Cadmium - ICP-Acid Digest	ug/g	89.40	<5.00e-03	4.440	4.640	4.540	4.41	83.20	17.80	n/a
S961003988	A	Cerium - ICP-Acid Digest	ug/g	93.60	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	83.00	35.60	n/a
S961003988	A	Cobalt - ICP-Acid Digest	ug/g	91.80	<2.00e-02	< 7.130	<7.11e0	n/a	n/a	84.60	7.120	n/a
S961003988	A	Chromium - ICP-Acid Digest	ug/g	90.40	<1.00e-02	1.17e+03	1.12e+03	1.14e+03	4.37	80.00	3.560	n/a
S961003988	A	Copper - ICP-Acid Digest	ug/g	89.80	<1.00e-02	4.340	4.430	4.385	2.05	80.80	3.560	n/a
S961003988	A	Iron - ICP-Acid Digest	ug/g	93.80	1.20e-01	1.83e+02	1.75e1	179.0	4.47	87.80	17.80	n/a
S961003988	A	Potassium - ICP-Acid Digest	ug/g	91.00	<5.00e-02	2.94e+03	2.90e+03	2.92e+03	1.37	98.00	17.80	n/a
S961003988	A	Lanthanum - ICP-Acid Digest	ug/g	91.60	<1.78e1	< 17.80	<1.78e1	n/a	n/a	85.60	17.80	n/a
S961003988	A	Magnesium - ICP-Acid Digest	ug/g	93.40	<1.00e-01	64.30	62.80	63.55	2.36	86.80	3.560	n/a
S961003988	A	Lithium - ICP-Acid Digest	ug/g	88.40	<1.00e-02	9.400	8.890	9.145	5.58	79.20	35.60	n/a
S961003988	A	Manganese - ICP-Acid Digest	ug/g	90.60	<5.00e-02	43.80	43.70	43.75	0.23	82.40	3.560	n/a
S961003988	A	Molybdenum - ICP-Acid Digest	ug/g	90.60	<5.00e-02	1.80e+05	1.71e+05	1.76e+05	5.13	-1.870e3	35.60	n/a
S961003988	A	Neodymium - ICP-Acid Digest	ug/g	92.60	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	84.20	35.60	n/a
S961003988	A	Nickel - ICP-Acid Digest	ug/g	92.20	<2.00e-02	24.60	22.50	23.55	8.92	84.80	7.120	n/a
S961003988	A	Phosphorus - ICP-Acid Digest	ug/g	93.00	<2.00e-01	4.55e+02	4.55e1	455.0	0.00	76.80	71.20	n/a
S961003988	A	Lead - ICP-Acid Digest	ug/g	87.80	<1.00e-01	36.40	<3.56e1	n/a	n/a	82.20	35.60	n/a
S961003988	A	Sulfur - ICP-Acid Digest	ug/g	91.60	<1.00e-01	2.82e+03	2.71e+03	2.76e+03	3.98	75.00	35.60	n/a
S961003988	A	Antimony - ICP-Acid Digest	ug/g	89.60	<6.00e-02	< 21.40	<2.13e1	n/a	n/a	84.20	21.40	n/a
S961003988	A	Selenium - ICP-Acid Digest	ug/g	85.20	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	79.80	35.60	n/a
S961003988	A	Silicon - ICP-Acid Digest	ug/g	118.2	2.57e-01	1.36e+02	135.0	135.5	0.74	88.80	17.80	n/a
S961003988	A	Samarium - ICP-Acid Digest	ug/g	90.40	<1.00e-01	< 35.60	<3.56e1	n/a	n/a	83.00	35.60	n/a
S961003988	A	Strontium - ICP-Acid Digest	ug/g	90.80	<1.00e-02	< 3.560	<3.56e0	n/a	n/a	82.40	3.560	n/a
S961003988	A	Titanium - ICP-Acid Digest	ug/g	83.60	<1.00e-02	< 3.560	<3.56e0	n/a	n/a	76.00	3.560	n/a
S961003988	A	Thallium - ICP-Acid Digest	ug/g	84.80	<5.00e-01	< 71.30	<7.11e1	n/a	n/a	78.40	71.20	n/a
S961003988	A	Uranium - ICP-Acid Digest	ug/g	88.60	<5.00e-01	<1.78e2	<1.78e2	n/a	n/a	85.50	17.80	n/a
S961003988	A	Vanadium - ICP-Acid Digest	ug/g	89.80	<5.00e-02	< 17.80	<1.78e1	n/a	n/a	82.20	17.80	n/a
S961003988	A	Zinc - ICP-Acid Digest	ug/g	89.40	1.06e-01	38.00	47.40	42.70	22.0	83.60	3.560	n/a
S961003988	A	Zirconium - ICP-Acid Digest	ug/g	102.4	<2.00e-02	5.930	5.550	5.740	6.62	83.20	3.560	n/a
S961003990	A	Acetate by IC-DIONEX 4000/4500	ug/g	101.4	<2.00e-02	8.23e+02	779.0	776.2	12.1	121.8	82.49	n/a
S961003990	A	Formate by IC-DIONEX 4000/4500	ug/g	102.4	<2.00e-02	1.26e+03	1.17e+03	1.22e+03	7.41	105.0	86.60	n/a
S961003990	A	Fluoride by IC-DIONEX 4000/4500	ug/g	102.4	<1.30e-02	1.15e+03	1.38e+03	1.26e+03	18.2	68.64	36.58	n/a
S961003990	A	Chloride by IC-DIONEX 4000/4500	ug/g	102.4	1.70e-02	3.94e+03	5.28e+03	4.61e+03	29.1	126.2	47.85	n/a
S961003990	A	Nitrite by IC-DIONEX 4000/4500	ug/g	98.12	<1.07e-01	4.81e+04	6.53e+04	5.67e+04	30.3	152.6	301.1	n/a
S961003990	A	Bromide by Ion Chromatograph	ug/g	98.47	<1.26e-01	5.75e+02	625.0	600.2	8.33	92.02	354.5	n/a
S961003990	A	Nitrate by IC-DIONEX 4000/4500	ug/g	101.7	1.57e-01	1.86e+05	1.86e+05	1.51e+05	46.4	176.8	393.6	n/a
S961003990	A	Phosphate by IC-DIONEX 4000/4500	ug/g	100.0	<1.36e-01	1.63e+03	996.0	7.30e+03	48.3	99.27	334.6	n/a
S961003990	A	Sulfate by IC-DIONEX 4000/4500	ug/g	101.4	<1.36e-01	6.04e+03	8.56e+03	7.30e+03	34.5	106.3	382.5	n/a
S961003990	A	Oxalate by IC-DIONEX 4000/4500	ug/g	95.26	<1.05e-01	2.89e+03	3.90e+03	3.39e+03	29.7	96.58	295.5	n/a

Drainable Liquid; Drainable Liquid

Sample/	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S961003980		DSC Exotherm DRY, Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S961003980		DSC Exotherm on Perkin Elmer	Joules/g	104.4	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a	n/a
S961003980		Specific Gravity	Sp. G.	99.14	n/a	1.412	1.408	1.410	0.28	n/a	1.00e-02	n/a	n/a
S961003980		% Water by TGA on Perkin Elmer	%	99.34	n/a	53.90	54.21	54.05	0.57	n/a	n/a	n/a	n/a
S961003980	D	Silver- ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	15.70	15.50	15.60	1.28	87.80	6.010	n/a	n/a
S961003980	D	Aluminum- ICP-Acid Dil.	ug/mL	98.20	<5.00e-02	2.74e+04	2.70e+04	2.72e+04	1.47	-2.20e1	30.10	n/a	n/a
S961003980	D	Arsenic- ICP-Acid Dil.	ug/mL	104.6	<1.00e-02	62.60	<6.010	61.25	4.41	95.50	30.10	n/a	n/a
S961003980	D	Boron- ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	<3.010	<3.010	n/a	n/a	96.10	30.10	n/a	n/a
S961003980	D	Barium- ICP-Acid Dil.	ug/mL	100.00	<5.00e-02	<3.010	<3.010	n/a	n/a	96.10	30.10	n/a	n/a
S961003980	D	Beryllium- ICP-Acid Dil.	ug/mL	102.6	<5.00e-03	<3.000	<3.000	n/a	n/a	99.20	3.000	n/a	n/a
S961003980	D	Bismuth- ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	<6.010	<6.010	n/a	n/a	99.90	60.10	n/a	n/a
S961003980	D	Calcium- ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	<6.010	<6.010	n/a	n/a	100.0	60.10	n/a	n/a
S961003980	D	Cadmium- ICP-Acid Dil.	ug/mL	100.6	<5.00e-03	<3.000	<3.000	n/a	n/a	98.60	3.000	n/a	n/a
S961003980	D	Cerium- ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	<6.010	<6.010	n/a	n/a	98.50	60.10	n/a	n/a
S961003980	D	Cobalt- ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	12.00	<1.20e1	n/a	n/a	98.20	12.00	n/a	n/a
S961003980	D	Copper- ICP-Acid Dil.	ug/mL	100.8	<1.00e-02	1.57e+02	155.0	156.0	1.28	98.20	6.010	n/a	n/a
S961003980	D	Chromium- ICP-Acid Dil.	ug/mL	103.6	<1.00e-02	<6.010	<6.010	n/a	n/a	99.10	6.010	n/a	n/a
S961003980	D	Iron- ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	<3.010	<3.010	n/a	n/a	98.80	30.10	n/a	n/a
S961003980	D	Potassium- ICP-Acid Dil.	ug/mL	100.2	<5.00e-01	5.08e+03	4.97e+03	5.02e+03	2.19	85.60	300.10	n/a	n/a
S961003980	D	Lanthanum- ICP-Acid Dil.	ug/mL	101.4	<5.00e-02	<3.010	<3.010	n/a	n/a	99.30	30.10	n/a	n/a
S961003980	D	Lithium- ICP-Acid Dil.	ug/mL	102.0	<1.00e-02	<6.010	<6.010	n/a	n/a	100.0	6.010	n/a	n/a
S961003980	D	Magnesium- ICP-Acid Dil.	ug/mL	96.60	<1.00e-01	<6.010	<6.010	n/a	n/a	92.20	60.10	n/a	n/a
S961003980	D	Manganese- ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	<6.010	<6.010	n/a	n/a	93.70	6.010	n/a	n/a
S961003980	D	Molybdenum- ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	72.80	71.00	71.90	2.50	98.70	30.10	n/a	n/a
S961003980	D	Sodium- ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	2.23e+05	2.19e+05	2.21e+05	1.81	-8.94e2	60.10	n/a	n/a
S961003980	D	Neodymium- ICP-Acid Dil.	ug/mL	101.4	<1.00e-01	<6.010	<6.010	n/a	n/a	99.00	60.10	n/a	n/a
S961003980	D	Nickel- ICP-Acid Dil.	ug/mL	99.80	<2.00e-02	12.00	<1.20e1	n/a	n/a	96.70	12.00	n/a	n/a
S961003980	D	Phosphorus- ICP-Acid Dil.	ug/mL	100.4	<2.00e-01	3.87e+02	379.0	383.0	2.09	92.50	120.10	n/a	n/a
S961003980	D	Lead- ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	<6.010	<6.010	n/a	n/a	103.0	60.10	n/a	n/a
S961003980	D	Sulfur- ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	1.33e+03	1.31e+03	1.32e+03	1.52	91.00	60.10	n/a	n/a
S961003980	D	Antimony- ICP-Acid Dil.	ug/mL	95.80	<6.00e-02	<36.10	<3.61e1	n/a	n/a	93.80	36.10	n/a	n/a
S961003980	D	Selenium- ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	<6.010	<6.010	n/a	n/a	99.50	60.10	n/a	n/a
S961003980	D	Silicon- ICP-Acid Dil.	ug/mL	98.80	<5.00e-02	1.96e+02	191.0	193.5	2.58	98.40	30.10	n/a	n/a
S961003980	D	Samarium- ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	<6.010	<6.010	n/a	n/a	96.20	60.10	n/a	n/a
S961003980	D	Strontium- ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	<6.010	<6.010	n/a	n/a	96.00	6.010	n/a	n/a
S961003980	D	Titanium- ICP-Acid Dil.	ug/mL	96.20	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	95.00	6.010	n/a	n/a
S961003980	D	Thallium- ICP-Acid Dil.	ug/mL	97.60	<5.00e-01	<3.00e+02	<3.00e2	n/a	n/a	94.20	120.10	n/a	n/a
S961003980	D	Vanadium- ICP-Acid Dil.	ug/mL	96.10	<5.00e-02	<3.00e+02	<3.00e2	n/a	n/a	96.50	300.10	n/a	n/a
S961003980	D	Uranium- ICP-Acid Dil.	ug/mL	100.2	<5.00e-01	<30.10	<3.01e1	n/a	n/a	97.80	30.10	n/a	n/a
S961003980	D	Zinc- ICP-Acid Dil.	ug/mL	101.2	<1.00e-02	39.30	38.30	38.80	2.58	99.30	6.010	n/a	n/a
S961003980	D	Zirconium- ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	<6.010	<6.010	n/a	n/a	97.20	6.010	n/a	n/a
S961003980		Acetate by IC-Dionex 4000/4500	ug/mL	100.8	<2.10e-02	1.31e+03	1.27e+03	1.29e+03	6.30	91.13	4.020	n/a	n/a
S961003980		Formate by Ion Chromatograph	ug/mL	95.90	<2.10e-02	1.49e+03	1.37e+03	1.43e+03	8.39	80.92	4.221	n/a	n/a
S961003980		Fluoride- IC-Dionex 4000/4500	ug/mL	103.2	<1.00e-02	<53.83	<5.38e1	n/a	n/a	n/a	53.83	n/a	n/a
S961003980		Chloride- IC-Dionex 4000/4500	ug/mL	99.49	<1.00e-02	9.31e+03	9.14e+03	9.23e+03	1.84	n/a	70.40	n/a	n/a
S961003980		Nitrite- IC - Dionex 4000/4500	ug/mL	98.35	<1.07e-01	9.48e+04	9.39e+04	9.43e+04	0.95	n/a	441.1	n/a	n/a
S961003980		Bromide by Ion Chromatograph	ug/mL	98.30	<1.26e-01	5.90e+03	5.92e+03	5.92e+03	0.84	n/a	521.8	n/a	n/a
S961003980		Nitrate by IC-Dionex 4000/4500	ug/mL	99.50	1.80e-01	1.74e+05	1.73e+05	1.74e+05	0.58	n/a	579.7	n/a	n/a

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003980		Phosphate-IC-Dionex 4000/4500	ug/mL	97.43	<1.19e-01	7.45e+02	532.0	638.5	33.4	n/a	492.4	n/a
S96T003980		Sulfate by IC-Dionex 4000/4500	ug/mL	99.53	<1.36e-01	2.84e+03	2.86e+03	2.85e+03	0.70	n/a	562.8	n/a
S96T003980		Oxalate by IC-Dionex 4000/4500	ug/mL	100.6	<1.05e-01	<4.35e+02	<4.35e+02	n/a	n/a	n/a	434.8	n/a
S96T003980		Alpha In Liquid Samples	ucf/mL	93.75	<2.94e-02	<1.86e-02	<2.22E-2	n/a	n/a	n/a	3.90e-02	5.00E+02

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: U Upper Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003973		Bulk Density of Sample	g/mL		n/a	1.580	n/a	n/a	n/a	n/a	5.00e-01	n/a
S961003983		DSC Exotherm using Wetter	Joules/g	102.6	n/a	39.10	48.60	43.85	21.7	n/a	n/a	n/a
S961003983		DSC Exotherm Dry	Joules/g Dry	n/a	n/a	69.40	86.30	77.85	21.7	n/a	n/a	n/a
S961003983		% Water by TGA using Mettler	%	99.53	n/a	44.17	43.15	43.66	2.34	n/a	n/a	n/a
S961003985		F Silver - ICP-Fusion	ug/g	100.6	<1.00e-02	1.90e+04	2.03e+04	1.96e+04	6.62	n/a	82.20	200.0
S961003985		F Silver - ICP-Fusion	ug/g	98.80	<1.00e-02	1.90e+04	2.03e+04	1.96e+04	6.62	n/a	82.20	200.0
S961003986		F Aluminum - ICP-Fusion	ug/g	104.2	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	104.0	1.00e+03	n/a
S961003986		F Arsenic - ICP-Fusion	ug/g	102.8	<1.00e-02	<9.99e+02	<1.02e3	n/a	n/a	103.0	2.00e+03	n/a
S961003986		F Boron - ICP-Fusion	ug/g	99.80	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	99.10	1.00e+03	n/a
S961003986		F Barium - ICP-Fusion	ug/g	102.8	<5.00e-03	<2.00e+03	<2.03e3	n/a	n/a	102.0	100.0	n/a
S961003986		F Beryllium - ICP-Fusion	ug/g	103.2	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	102.0	2.00e+03	n/a
S961003986		F Bismuth - ICP-Fusion	ug/g	101.2	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	101.0	100.0	n/a
S961003986		F Calcium - ICP-Fusion	ug/g	100.6	<5.00e-03	<2.00e+03	<2.03e3	n/a	n/a	101.0	2.00e+03	n/a
S961003986		F Cadmium - ICP-Fusion	ug/g	100.6	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	101.0	2.00e+03	n/a
S961003986		F Cerium - ICP-Fusion	ug/g	103.0	<2.00e-02	<3.99e+02	<4.06e2	n/a	n/a	102.0	200.0	n/a
S961003986		F Cobalt - ICP-Fusion	ug/g	101.8	<1.00e-02	<2.00e+02	<2.03e2	2.66e+03	4.88	101.0	200.0	n/a
S961003986		F Chromium - ICP-Fusion	ug/g	105.6	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	107.0	200.0	n/a
S961003986		F Copper - ICP-Fusion	ug/g	101.6	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	102.0	1.00e+03	n/a
S961003986		F Iodine - ICP-Fusion	ug/g	93.00	<5.00e-01	6.12e+06	n/a	n/a	n/a	n/a	1.00e+04	n/a
S961003986		F Potassium - ICP-Fusion	ug/g	101.6	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	102.0	1.00e+04	n/a
S961003986		F Lanthanum - ICP-Fusion	ug/g	102.2	<1.00e-02	<2.00e+03	<2.03e3	n/a	n/a	104.0	200.0	n/a
S961003986		F Lithium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	99.10	2.00e+03	n/a
S961003986		F Magnesium - ICP-Fusion	ug/g	99.20	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	96.60	200.0	n/a
S961003986		F Manganese - ICP-Fusion	ug/g	103.4	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	103.0	1.00e+03	n/a
S961003986		F Molybdenum - ICP-Fusion	ug/g	99.40	1.640	2.38e+05	2.35e+05	2.36e+05	1.27	99.90	2.00e+03	n/a
S961003986		F Nickel - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	103.0	2.00e+03	n/a
S961003986		F Neodymium - ICP-Fusion	ug/g	100.8	7.80e-02	<3.99e+02	<4.06e2	n/a	n/a	102.0	1.00e+03	n/a
S961003986		F Nickel - ICP-Fusion	ug/g	106.8	<2.00e-01	5.14e+03	4.57e+03	4.86e+03	11.7	112.0	3.99e+03	n/a
S961003986		F Phosphorus - ICP-Fusion	ug/g	101.0	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	98.80	2.00e+03	n/a
S961003986		F Lead - ICP-Fusion	ug/g	102.6	<1.00e-01	2.52e+03	2.53e+03	2.52e+03	0.40	106.0	2.00e+03	n/a
S961003986		F Sulfur - ICP-Fusion	ug/g	98.20	<6.00e-02	<1.20e+03	<1.22e3	n/a	n/a	99.30	1.20e+03	n/a
S961003986		F Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.00e+03	<2.03e3	n/a	n/a	99.30	2.00e+03	n/a
S961003986		F Selenium - ICP-Fusion	ug/g	96.80	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	99.70	2.00e+03	n/a
S961003986		F Silicon - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.00e+03	<2.03e3	n/a	n/a	99.70	2.00e+03	n/a
S961003986		F Samarium - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.00e+03	<2.03e3	n/a	n/a	99.70	2.00e+03	n/a
S961003986		F Strontium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	100.0	3.99e+03	n/a
S961003986		F Tantalum - ICP-Fusion	ug/g	97.70	<2.00e-01	<3.99e+03	<4.06e3	n/a	n/a	99.00	1.00e+04	n/a
S961003986		F Thallium - ICP-Fusion	ug/g	97.40	<5.00e-01	<9.99e+03	<1.02e4	n/a	n/a	102.0	1.00e+04	n/a
S961003986		F Uranium - ICP-Fusion	ug/g	102.2	<5.00e-02	<9.99e+02	<1.02e3	n/a	n/a	103.0	200.0	n/a
S961003986		F Vanadium - ICP-Fusion	ug/g	103.6	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	101.0	2.00e+03	n/a
S961003986		F Zinc - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.00e+02	<2.03e2	n/a	n/a	101.0	4.550	n/a
S961003986		F Zirconium - ICP-Fusion	ug/g	95.20	<1.00e-02	19.10	16.20	17.65	16.4	n/a	4.550	n/a
S961003989		A Silver - ICP-Acid Digest	ug/g	97.60	1.03e-01	1.93e+04	2.08e+04	2.00e+04	7.45	n/a	23.20	n/a
S961003989		A Aluminum - ICP-Acid Digest	ug/g	97.60	1.03e-01	1.93e+04	2.08e+04	2.00e+04	7.45	n/a	23.20	n/a

A-0002-3

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003989	A	Arsenic - ICP-Acid Digest	ug/g	97.60	< 1.00e-01	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Boron - ICP-Acid Digest	ug/g	106.2	5.84e-01	82.30	97.50	89.90	16.9	n/a	23.20	n/a
S96T003989	A	Barium - ICP-Acid Digest	ug/g	99.40	< 5.00e-02	2.30	< 2.3161	n/a	n/a	n/a	23.20	n/a
S96T003989	A	Beryllium - ICP-Acid Digest	ug/g	99.20	< 5.00e-03	< 2.330	< 2.3160	n/a	n/a	n/a	2.330	n/a
S96T003989	A	Bismuth - ICP-Acid Digest	ug/g	93.80	< 1.00e-01	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Calcium - ICP-Acid Digest	ug/g	193.6	2.550	1.00e+03	1.10e+03	18.2	n/a	n/a	46.50	n/a
S96T003989	A	Cadmium - ICP-Acid Digest	ug/g	92.80	< 5.00e-03	11.40	9.850	10.62	14.6	n/a	2.330	n/a
S96T003989	A	Cerium - ICP-Acid Digest	ug/g	97.40	< 1.00e-01	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Cobalt - ICP-Acid Digest	ug/g	94.80	< 2.00e-02	< 9.300	< 9.2460	n/a	n/a	n/a	9.300	n/a
S96T003989	A	Chromium - ICP-Acid Digest	ug/g	94.00	< 1.00e-02	2.36e+03	2.03e+03	15.0	n/a	n/a	4.650	n/a
S96T003989	A	Copper - ICP-Acid Digest	ug/g	94.00	1.90e-02	6.280	4.790	5.535	26.9	n/a	4.650	n/a
S96T003989	A	Iron - ICP-Acid Digest	ug/g	94.60	< 5.00e-02	1.42e+02	134.0	133.0	13.5	n/a	23.20	n/a
S96T003989	A	Potassium - ICP-Acid Digest	ug/g	97.40	< 5.00e-01	3.03e+03	3.35e+03	3.19e+03	10.0	n/a	23.20	n/a
S96T003989	A	Lanthanum - ICP-Acid Digest	ug/g	97.40	< 5.00e-02	< 23.30	< 2.3161	n/a	n/a	n/a	23.20	n/a
S96T003989	A	Lithium - ICP-Acid Digest	ug/g	99.00	< 1.00e-02	8.850	12.20	10.52	31.8	n/a	4.650	n/a
S96T003989	A	Magnesium - ICP-Acid Digest	ug/g	91.60	< 1.00e-02	48.80	42.30	45.55	14.3	n/a	4.650	n/a
S96T003989	A	Manganese - ICP-Acid Digest	ug/g	92.40	< 5.00e-02	46.80	51.90	49.35	10.3	n/a	23.20	n/a
S96T003989	A	Molybdenum - ICP-Acid Digest	ug/g	131.2	1.150	1.88e+05	1.78e+05	1.83e+05	5.46	n/a	46.50	n/a
S96T003989	A	Sodium - ICP-Acid Digest	ug/g	96.60	< 1.00e-01	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Neodymium - ICP-Acid Digest	ug/g	95.80	< 2.00e-02	35.50	28.60	32.05	21.5	n/a	9.300	n/a
S96T003989	A	Nickel - ICP-Acid Digest	ug/g	96.00	< 2.00e-01	5.30e+03	3.85e+03	4.58e+03	31.7	n/a	93.00	n/a
S96T003989	A	Phosphorus - ICP-Acid Digest	ug/g	90.80	< 1.00e-01	47.90	51.90	49.90	8.02	n/a	46.50	n/a
S96T003989	A	Lead - ICP-Acid Digest	ug/g	97.20	1.15e-01	2.37e+03	2.19e+03	2.28e+03	7.89	n/a	46.50	n/a
S96T003989	A	Sulfur - ICP-Acid Digest	ug/g	93.00	< 6.00e-02	< 27.90	< 2.7761	n/a	n/a	n/a	28.00	n/a
S96T003989	A	Selenium - ICP-Acid Digest	ug/g	90.00	< 1.00e-01	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Silicon - ICP-Acid Digest	ug/g	126.4	1.29e-01	1.67e+02	136.0	151.5	20.5	n/a	23.20	n/a
S96T003989	A	Samarium - ICP-Acid Digest	ug/g	94.60	< 1.00e-02	< 46.50	< 46.621	n/a	n/a	n/a	46.50	n/a
S96T003989	A	Strontium - ICP-Acid Digest	ug/g	94.60	< 1.00e-02	< 4.650	< 4.6260	n/a	n/a	n/a	4.650	n/a
S96T003989	A	Titanium - ICP-Acid Digest	ug/g	86.80	< 2.00e-01	< 93.00	< 9.2461	n/a	n/a	n/a	93.00	n/a
S96T003989	A	Thallium - ICP-Acid Digest	ug/g	88.80	< 5.00e-01	< 2.33e+02	< 2.3162	n/a	n/a	n/a	232.0	n/a
S96T003989	A	Uranium - ICP-Acid Digest	ug/g	92.40	< 5.00e-02	< 23.30	< 2.3161	n/a	n/a	n/a	23.20	n/a
S96T003989	A	Vanadium - ICP-Acid Digest	ug/g	93.00	< 4.00e-02	49.50	51.50	50.50	3.96	n/a	4.650	n/a
S96T003989	A	Zinc - ICP-Acid Digest	ug/g	92.40	< 6.00e-02	6.950	< 46.621	n/a	n/a	n/a	4.650	n/a
S96T003989	A	Zirconium - ICP-Acid Digest	ug/g	95.00	< 1.00e-02	7.20e+02	591.0	655.3	19.7	n/a	61.29	n/a
S96T003992	W	Acetate by IC-DIONEX 4000/4500	ug/g	105.6	< 2.00e-02	1.07e+03	1.05e+03	1.05e+03	3.81	n/a	43.44	n/a
S96T003992	W	Formate by IC-DIONEX 4000/4500	ug/g	95.83	< 2.00e-02	2.68e+03	2.50e+03	2.59e+03	6.95	n/a	63.57	n/a
S96T003992	W	Fluoride-IC-Dionex 4000/4500	ug/g	102.5	1.30e-02	4.90e+03	4.78e+03	4.84e+03	2.48	n/a	83.12	n/a
S96T003992	W	Chloride-IC-Dionex 4000/4500	ug/g	95.82	< 1.00e-02	6.05e+04	5.78e+04	5.88e+04	5.96	n/a	523.1	n/a
S96T003992	W	Nitrite-IC - Dionex 4000/4500	ug/g	96.49	< 1.07e-01	6.10e+02	< 6.3062	n/a	n/a	n/a	615.9	n/a
S96T003992	W	Bromide by Ion Chromatograph	ug/g	94.74	< 1.26e-01	1.58e+05	1.69e+05	1.64e+05	6.73	n/a	684.1	n/a
S96T003992	W	Nitrate by IC-Dionex 4000/4500	ug/g	95.12	1.86e-01	1.21e+04	1.04e+04	1.12e+04	15.1	n/a	581.4	n/a
S96T003992	W	Phosphate-IC-Dionex 4000/4500	ug/g	95.41	< 1.19e-01	5.97e+03	5.69e+03	5.83e+03	4.80	n/a	96.67	n/a
S96T003992	W	Sulfate by IC-Dionex 4000/4500	ug/g	96.04	< 1.36e-01	6.06e+03	5.58e+03	5.82e+03	8.25	n/a	513.2	n/a
S96T003992	W	Oxalate by IC-Dionex 4000/4500	ug/g	100.4	< 1.05e-01	6.06e+03	5.58e+03	5.82e+03	8.25	n/a	513.2	n/a

WHO-SD-WM-IDP-199, REV. 1

WHO-SD-WV-IDP-199, REV. 1

L Lower Half of Segment: L Lower Half of Segment

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003972		Bulk Density of Sample	g/ml	n/a	n/a	1.600	n/a	n/a	n/a	n/a	5.00e-01	n/a
S96T003982		DSC Exotherm Dry Calculated	Joules/g	102.6	n/a	57.50	94.80	76.15	49.0	n/a	n/a	n/a
S96T003982		% Water by TGA using Mettler	%	99.53	n/a	1.06e-02	174.0	140.0	48.6	n/a	n/a	n/a
S96T003982		F Silver - ICP-Fusion	ug/g	100.6	<1.00e-02	45.87	45.21	45.54	1.45	n/a	201.0	n/a
S96T003985		F Aluminum - ICP-Fusion	ug/g	98.80	<1.00e-02	1.38e-04	1.43e+04	1.40e+04	3.56	99.10	1.01e+03	n/a
S96T003985		F Arsenic - ICP-Fusion	ug/g	104.2	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	105.0	2.01e+03	n/a
S96T003985		F Boron - ICP-Fusion	ug/g	99.80	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	101.0	1.01e+03	n/a
S96T003985		F Barium - ICP-Fusion	ug/g	99.80	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	98.10	1.01e+03	n/a
S96T003985		F Beryllium - ICP-Fusion	ug/g	102.8	<1.00e-03	<1.01e-02	<1.01e-02	n/a	n/a	102.0	101.0	n/a
S96T003985		F Bismuth - ICP-Fusion	ug/g	103.2	<1.00e-01	<2.01e-03	<2.01e-03	n/a	n/a	100.0	2.01e+03	n/a
S96T003985		F Calcium - ICP-Fusion	ug/g	101.1	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	103.0	2.01e+03	n/a
S96T003985		F Cadmium - ICP-Fusion	ug/g	101.6	<5.00e-03	<2.01e-02	<2.01e-02	n/a	n/a	99.80	2.01e+03	n/a
S96T003985		F Cerium - ICP-Fusion	ug/g	103.0	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	102.0	403.0	n/a
S96T003985		F Cobalt - ICP-Fusion	ug/g	105.6	<1.00e-02	5.68e-03	5.68e-03	5.58e-03	3.40	100.0	201.0	n/a
S96T003985		F Chromium - ICP-Fusion	ug/g	105.6	<1.00e-02	<2.01e+03	<2.01e-03	n/a	n/a	106.0	201.0	n/a
S96T003985		F Copper - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	102.0	1.01e+03	n/a
S96T003985		F Iron - ICP-Fusion	ug/g	93.00	<5.00e-01	5.54e+06	n/a	n/a	n/a	101.0	1.01e+04	n/a
S96T003985		F Potassium - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	102.0	201.0	n/a
S96T003985		F Lanthanum - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	97.70	2.01e+03	n/a
S96T003985		F Magnesium - ICP-Fusion	ug/g	99.20	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	96.60	201.0	n/a
S96T003985		F Manganese - ICP-Fusion	ug/g	103.4	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	103.0	1.01e+03	n/a
S96T003985		F Molybdenum - ICP-Fusion	ug/g	99.40	<1.00e-01	2.13e+05	2.14e+05	2.14e+05	0.47	101.0	2.01e+03	n/a
S96T003985		F Sodium - ICP-Fusion	ug/g	102.0	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	102.0	2.01e+03	n/a
S96T003985		F Neodymium - ICP-Fusion	ug/g	100.8	<7.80e-02	6.24e+03	1.65e+03	3.94e+03	116	98.70	403.0	n/a
S96T003985		F Nickel - ICP-Fusion	ug/g	106.8	<2.00e-01	3.57e+04	3.31e+04	3.44e+04	7.56	101.0	4.03e+03	n/a
S96T003985		F Phosphorus - ICP-Fusion	ug/g	101.0	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	99.20	2.01e+03	n/a
S96T003985		F Lead - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	98.90	2.01e+03	n/a
S96T003985		F Sulfur - ICP-Fusion	ug/g	98.20	<1.00e-02	<1.21e+03	<1.21e-03	n/a	n/a	96.00	1.21e+03	n/a
S96T003985		F Antimony - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	98.30	2.01e+03	n/a
S96T003985		F Selenium - ICP-Fusion	ug/g	96.80	<5.00e-02	<1.01e+03	<1.01e-03	n/a	n/a	100.0	1.01e+03	n/a
S96T003985		F Silicon - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.01e+03	<2.01e-03	n/a	n/a	97.10	2.01e+03	n/a
S96T003985		F Samarium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.01e+03	<2.01e-03	n/a	n/a	98.60	201.0	n/a
S96T003985		F Strontium - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.01e+03	<2.01e-03	n/a	n/a	99.30	201.0	n/a
S96T003985		F Tantalum - ICP-Fusion	ug/g	97.40	<2.00e-01	4.03e+03	4.03e-03	n/a	n/a	96.10	4.03e+03	n/a
S96T003985		F Thallium - ICP-Fusion	ug/g	97.70	<5.00e-01	<1.01e+04	<1.01e-04	n/a	n/a	93.00	1.01e+04	n/a
S96T003985		F Vanadium - ICP-Fusion	ug/g	103.6	<1.00e-02	<2.01e+03	<2.01e-03	n/a	n/a	101.0	1.01e+03	n/a
S96T003985		F Zinc - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.01e+03	<2.01e-03	n/a	n/a	103.0	201.0	n/a
S96T003985		F Zirconium - ICP-Fusion	ug/g	96.88	<2.99e-03	1.33e+02	480.0	426.5	25.1	100.0	201.0	n/a
S96T003987		F Alpha of Digested Solid	ug/g	95.20	<1.00e-02	98.70	107.0	102.8	8.07	88.86	4.00e-03	1.08e+01
S96T003987		A Silver - ICP-Acid Digest	ug/g	97.60	<1.00e-01	1.52e+04	1.58e+04	8.20	-8.64e-01	79.10	24.40	n/a
S96T003987		A Aluminum - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<48.70	<4.89e1	n/a	n/a	93.80	48.80	n/a
S96T003987		A Arsenic - ICP-Acid Digest	ug/g	106.2	<5.84e-01	1.44e+02	109.0	126.5	27.7	85.20	24.40	n/a
S96T003987		A Boron - ICP-Acid Digest	ug/g	95.40	<5.00e-02	<24.40	<2.44e1	n/a	n/a	91.40	24.40	n/a
S96T003987		A Barium - ICP-Acid Digest	ug/g	99.20	<5.00e-03	<2.440	<2.44e0	n/a	n/a	93.20	2.440	n/a
S96T003987		A Beryllium - ICP-Acid Digest	ug/g	99.20	<5.00e-03	<2.440	<2.44e0	n/a	n/a	93.20	2.440	n/a

Sample#	RAW Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%	
S961003987	A Bismuth - ICP-Acid Digest	ug/g	93.80	< 1.00e-01	62.60	71.80	67.20	13.7	90.00	48.80	n/a	
S961003987	A Calcium - ICP-Acid Digest	ug/g	193.4	2.250	2.01e+03	2.27e+03	2.14e+03	12.1	38.80	48.80	n/a	
S961003987	A Cadmium - ICP-Acid Digest	ug/g	92.80	< 5.00e-03	36.90	40.10	38.50	8.31	89.20	2.440	n/a	
S961003987	A Cobalt - ICP-Acid Digest	ug/g	97.40	< 1.00e-01	< 4.70	< 4.89e1	n/a	n/a	93.00	48.80	n/a	
S961003987	A Chromium - ICP-Acid Digest	ug/g	94.80	< 2.00e-02	< 9.70	< 9.78e0	n/a	n/a	91.00	9.750	n/a	
S961003987	A Copper - ICP-Acid Digest	ug/g	94.00	< 1.00e-02	6.10e+03	6.51e+03	6.30e+03	6.50	45.80	4.880	n/a	
S961003987	A Iron - ICP-Acid Digest	ug/g	94.00	< 5.00e-02	6.820	18.10	12.46	90.5	91.40	4.880	n/a	
S961003987	A Potassium - ICP-Acid Digest	ug/g	97.40	< 5.00e-01	2.38e+02	2.59e+02	494.5	17.6	73.00	24.40	n/a	
S961003987	A Lanthanum - ICP-Acid Digest	ug/g	97.40	< 5.00e-02	< 24.40	< 2.44e1	n/a	n/a	93.80	24.40	n/a	
S961003987	A Lithium - ICP-Acid Digest	ug/g	99.00	< 1.00e-02	26.70	28.30	27.50	5.82	94.60	4.880	n/a	
S961003987	A Magnesium - ICP-Acid Digest	ug/g	94.40	< 1.00e-01	1.51e+02	165.0	158.0	8.86	83.80	48.80	n/a	
S961003987	A Manganese - ICP-Acid Digest	ug/g	91.60	< 1.00e-02	1.57e+02	165.0	158.0	8.86	85.00	4.880	n/a	
S961003987	A Molybdenum - ICP-Acid Digest	ug/g	92.40	< 5.00e-02	36.70	37.50	37.10	2.16	88.00	24.40	n/a	
S961003987	A Sodium - ICP-Acid Digest	ug/g	131.2	1.150	1.80e+05	1.81e+05	1.80e+05	0.55	-1.030e1	48.80	n/a	
S961003987	A Neodymium - ICP-Acid Digest	ug/g	96.60	< 1.00e-01	< 48.70	< 4.89e1	n/a	n/a	94.00	48.80	n/a	
S961003987	A Nickel - ICP-Acid Digest	ug/g	95.80	< 2.00e-02	65.20	67.20	66.20	3.02	91.80	9.750	n/a	
S961003987	A Phosphorus - ICP-Acid Digest	ug/g	96.00	< 2.00e-01	2.97e+04	2.70e+04	2.84e+04	9.52	80.20	97.50	n/a	
S961003987	A Lead - ICP-Acid Digest	ug/g	90.80	< 1.00e-01	1.15e+02	118.0	116.5	2.58	86.00	48.80	n/a	
S961003987	A Sulfur - ICP-Acid Digest	ug/g	97.20	1.15e-01	7.11e+02	754.0	732.5	5.87	82.60	48.80	n/a	
S961003987	A Antimony - ICP-Acid Digest	ug/g	93.00	< 6.00e-02	< 29.20	< 2.93e1	n/a	n/a	90.80	29.20	n/a	
S961003987	A Selenium - ICP-Acid Digest	ug/g	90.00	< 1.00e-01	< 48.70	< 4.89e1	n/a	n/a	86.20	48.80	n/a	
S961003987	A Silicon - ICP-Acid Digest	ug/g	126.4	1.29e-01	1.13e+02	109.0	111.0	3.60	79.00	24.40	n/a	
S961003987	A Samarium - ICP-Acid Digest	ug/g	94.60	< 1.00e-01	< 48.70	< 4.89e1	n/a	n/a	90.80	48.80	n/a	
S961003987	A Strontium - ICP-Acid Digest	ug/g	86.80	< 1.00e-02	< 4.870	< 4.89e0	9.720	9.395	7.13	89.80	4.880	n/a
S961003987	A Thallium - ICP-Acid Digest	ug/g	88.80	< 2.00e-01	< 97.40	< 9.78e1	n/a	n/a	80.60	4.880	n/a	
S961003987	A Uranium - ICP-Acid Digest	ug/g	92.40	< 5.00e-01	< 2.44e+02	< 2.44e2	n/a	n/a	94.30	244.0	n/a	
S961003987	A Vanadium - ICP-Acid Digest	ug/g	93.00	< 5.00e-02	< 24.40	< 2.44e1	n/a	n/a	90.00	24.40	n/a	
S961003987	A Zinc - ICP-Acid Digest	ug/g	92.40	4.60e-02	63.20	75.30	69.25	17.5	90.40	4.880	n/a	
S961003987	A Zirconium - ICP-Acid Digest	ug/g	95.00	< 1.00e-02	33.10	118.0	75.55	112	14.48	4.880	n/a	
S961003991	W Acetate by IC-DIONEX 4000/4500	ug/g	102.4	< 2.00e-02	< 82.50	< 8.30e1	n/a	n/a	116.9	82.51	n/a	
S961003991	W Formate by IC-DIONEX 4000/4500	ug/g	104.2	< 1.00e-02	9.85e+02	894.0	939.5	9.69	103.3	86.03	n/a	
S961003991	W Fluoride by IC-DIONEX 4000/4500	ug/g	95.76	< 1.00e-02	1.01e+04	8.65e+03	9.40e+03	15.5	94.24	53.63	n/a	
S961003991	W Chloride by IC-DIONEX 4000/4500	ug/g	99.37	< 1.00e-02	3.75e+03	3.48e+03	3.62e+03	7.47	103.5	70.12	n/a	
S961003991	W Nitrite by IC-DIONEX 4000/4500	ug/g	96.32	< 1.00e-01	4.34e+04	3.95e+04	4.13e+04	9.92	106.6	44.1	n/a	
S961003991	W Bromide by Ion Chromatograph	ug/g	94.40	< 1.26e-01	8.35e+02	793.0	814.0	5.16	98.30	519.8	n/a	
S961003991	W Nitrate by IC-DIONEX 4000/4500	ug/g	98.48	< 1.40e-01	8.88e+04	7.56e+04	8.22e+04	16.1	108.9	577.1	n/a	
S961003991	W Phosphate by IC-DIONEX 4000/4500	ug/g	94.68	< 1.19e-01	1.04e+05	1.02e+05	1.03e+05	1.94	103.0	490.4	n/a	
S961003991	W Sulfate by IC-DIONEX 4000/4500	ug/g	97.94	< 1.36e-01	1.34e+03	1.26e+03	1.30e+03	6.15	104.0	560.5	n/a	
S961003991	W Oxalate by IC-DIONEX 4000/4500	ug/g	96.79	< 1.05e-01	9.90e+03	8.73e+03	9.32e+03	12.6	103.8	433.0	n/a	

CORE NUMBER: 153
SEGMENT #: FB

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Field Blank

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003974	D	DSC Exotherm using Mettler	Joules/g	96.31	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003974	D	PSC Exotherm Dry, Calculated	Joules/g	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S96T003974	D	OH- by Pot. Titration	ug/mL	97.36	<42.00	< 42.00	<42	n/a	n/a	n/a	42.00	n/a
S96T003974	D	Specific Gravity	Sp.G.	99.14	n/a	9.82e-01	9.84e-01	9.84e-01	0.51	n/a	1.00e-02	n/a
S96T003974	D	% Water by TGA using Mettler	%	99.48	n/a	99.77	99.76	99.77	0.01	n/a	n/a	n/a
S96T003974	D	Silver-ICP-Acid Dil.	ug/mL	100.0	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	98.10	4.010	n/a
S96T003974	D	Aluminum-ICP-Acid Dil.	ug/mL	98.00	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	100.0	20.10	n/a
S96T003974	D	Arsenic-ICP-Acid Dil.	ug/mL	102.8	<4.00e-02	< 40.10	<4.01e1	n/a	n/a	101.0	40.10	n/a
S96T003974	D	Boron-ICP-Acid Dil.	ug/mL	99.40	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	97.20	20.10	n/a
S96T003974	D	Barium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	95.10	20.10	n/a
S96T003974	D	Beryllium-ICP-Acid Dil.	ug/mL	102.2	<5.00e-02	< 2.000	<2.00e0	n/a	n/a	99.60	2.000	n/a
S96T003974	D	Bismuth-ICP-Acid Dil.	ug/mL	101.8	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.90	40.10	n/a
S96T003974	D	Calcium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	< 72.50	<4.01e1	n/a	n/a	84.00	40.10	n/a
S96T003974	D	Cadmium-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	< 2.000	<2.00e0	n/a	n/a	101.0	2.000	n/a
S96T003974	D	Cerium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	98.20	40.10	n/a
S96T003974	D	Cobalt-ICP-Acid Dil.	ug/mL	101.4	<2.00e-02	< 8.020	<8.02e0	n/a	n/a	101.0	8.020	n/a
S96T003974	D	Chromium-ICP-Acid Dil.	ug/mL	103.4	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	101.0	4.010	n/a
S96T003974	D	Copper-ICP-Acid Dil.	ug/mL	101.0	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	101.0	20.10	n/a
S96T003974	D	Iron-ICP-Acid Dil.	ug/mL	96.60	<5.00e-02	<2.00e+02	<2.00e2	n/a	n/a	124.0	200.0	n/a
S96T003974	D	Potassium-ICP-Acid Dil.	ug/mL	102.2	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	96.60	20.10	n/a
S96T003974	D	Lanthanum-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	97.60	4.010	n/a
S96T003974	D	Lithium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	96.00	40.10	n/a
S96T003974	D	Magnesium-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	95.90	4.010	n/a
S96T003974	D	Manganese-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	99.70	20.10	n/a
S96T003974	D	Molybdenum-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	1.92e+02	203.0	197.5	5.57	95.60	40.10	n/a
S96T003974	D	Sodium-ICP-Acid Dil.	ug/mL	102.6	<2.00e-02	< 40.10	<4.01e1	n/a	n/a	99.20	40.10	n/a
S96T003974	D	Neodymium-ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	< 80.20	<8.02e1	n/a	n/a	100.0	8.020	n/a
S96T003974	D	Nickel-ICP-Acid Dil.	ug/mL	101.8	<2.00e-02	< 40.10	<4.01e1	n/a	n/a	100.0	40.10	n/a
S96T003974	D	Phosphorus-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	100.0	40.10	n/a
S96T003974	D	Lead-ICP-Acid Dil.	ug/mL	99.40	<1.00e-01	< 24.10	<4.01e1	n/a	n/a	83.70	40.10	n/a
S96T003974	D	Sulfur-ICP-Acid Dil.	ug/mL	96.20	<5.00e-02	< 68.10	<2.41e1	n/a	n/a	94.00	24.10	n/a
S96T003974	D	Antimony-ICP-Acid Dil.	ug/mL	97.20	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	92.60	40.10	n/a
S96T003974	D	Selenium-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	101.0	20.10	n/a
S96T003974	D	Silicon-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	< 40.10	<4.01e1	n/a	n/a	95.30	40.10	n/a
S96T003974	D	Samarium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	95.70	4.010	n/a
S96T003974	D	Strontium-ICP-Acid Dil.	ug/mL	98.00	<2.00e-02	< 4.010	<4.01e0	n/a	n/a	95.60	4.010	n/a
S96T003974	D	Tantalum-ICP-Acid Dil.	ug/mL	98.00	<1.00e-01	< 80.20	<8.02e1	n/a	n/a	95.80	80.20	n/a
S96T003974	D	Thallium-ICP-Acid Dil.	ug/mL	96.40	<5.00e-02	<2.00e+02	<2.00e2	n/a	n/a	94.00	200.0	n/a
S96T003974	D	Uranium-ICP-Acid Dil.	ug/mL	100.6	<5.00e-02	< 20.10	<2.01e1	n/a	n/a	98.70	20.10	n/a
S96T003974	D	Vanadium-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	< 39.60	<4.00	26.80	95.5	94.70	4.010	n/a
S96T003974	D	Zinc-ICP-Acid Dil.	ug/mL	101.4	<1.00e-02	< 4.010	<4.01e0	n/a	n/a	96.30	4.010	n/a
S96T003974	D	Zirconium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	< 1.020	<1.02e0	n/a	n/a	n/a	1.020	n/a
S96T003974	D	Acetate by IC-Dioxmex 4000/4500	ug/mL	100.8	<2.00e-02							

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003974		Formate by Ion Chromatograph	ug/mL	91.67	<2.10e-02	< 1.071	<1.0760	n/a	n/a	n/a	1.071	n/a
S96T003974		Fluoride-IC-Dionex 4000/4500	ug/mL	95.59	<1.30e-02	2.719	2.710	2.715	0.37	100.5	1.43e-01	n/a
S96T003974		Chloride-IC-Dionex 4000/4500	ug/mL	98.10	<1.70e-02	5.191	5.320	5.255	2.47	98.10	1.87e-01	n/a
S96T003974		Nitrite-IC - Dionex 4000/4500	ug/mL	97.24	<1.07e-01	50.89	50.80	50.84	0.20	106.1	1.177	n/a
S96T003974		Bromide by Ion Chromatograph	ug/mL	98.13	<1.26e-01	5.954	5.870	5.912	1.35	98.13	1.386	n/a
S96T003974		Nitrate by IC-Dionex 4000/4500	ug/mL	98.82	<1.49e-01	2.51e+02	245.0	248.1	2.42	96.97	1.540	n/a
S96T003974		Phosphate-IC-Dionex 4000/4500	ug/mL	95.96	<1.19e-01	12.19	11.70	11.95	4.18	99.82	1.308	n/a
S96T003974		Sulfate by IC-Dionex 4000/4500	ug/mL	97.94	<1.36e-01	4.896	4.760	4.828	2.90	99.05	1.495	n/a
S96T003974		Oxalate by IC-Dionex 4000/4500	ug/mL	99.22	<1.05e-01	3.986	4.100	4.043	2.72	101.5	1.155	n/a
S96T003974		Alpha in Liquid Samples	ug/mL	103.6	<1.35e-05	<1.75e-05	<1.15e-5	n/a	n/a	105.3	1.88e-05	5.00e+02
S96T004003		Beta in Liquid Samples	ug/mL	101.7	<5.54e-05	1.61e-01	1.75e-01	1.68e-01	8.33	99.34	6.16e-05	4.37E-01
S96T004003		Chromium by Phosphorescence	ug/mL	94.05	<2.20e-01	<0.419	<0.419	n/a	n/a	107.4	4.19e-01	n/a
S96T004003		U-Phosphorescence Inst. Error	% Inst Error	94.09	3.36e-01	3.38e-01	3.80e-01	3.59e-01	11.7	127.0	4.00e-03	1.70e+00
S96T004003		Technetium-99 Liq. Scint.	ug/mL	96.00	2.29e-05	1.700	1.240	1.470	1.24	n/a	n/a	n/a
S96T004003		Strontium-89/90 High Level	ug/mL	112.2	<8.88e-06	1.35e-04	1.32e-04	1.33e-04	2.25	n/a	8.01e-06	3.95e+00
S96T004003		Pu-239/240 by TRU-SPEC Resin	ug/mL	96.64	<3.55e-06	2.81e-03	2.90e-03	2.85e-03	3.15	n/a	1.04e-05	2.46e+00
S96T004003		Np237 by TTA Extraction	ug/mL	57.92	<2.06e-05	<3.61e-06	<3.43e-6	n/a	n/a	n/a	3.61e-06	1.00e+02
S96T004003		Iodine-129 Waste Tank Samples	ug/mL	86.51	<1.98e-05	<2.06e-05	<2.06e-5	n/a	n/a	n/a	4.32e-05	1.25e+02
S96T004003		Tritium By Lachat	ug/mL	104.2	1.31e-04	4.08e-05	<3.59e-5	n/a	n/a	96.68	2.17e-05	0.00e+00
S96T004003		R3 - Rel. % Count Error	% Ct. Error	1.01	4.600	5.190	<3.85e-5	5.530	5.53	n/a	3.85e-05	5.19e+00
S96T004003		Cobalt-60 by GEA	ug/mL	99.87	<3.42e-05	<2.55e-05	<3.41e-5	n/a	n/a	n/a	1.00e-04	n/a
S96T004003		Cesium-137 by GEA	ug/mL	96.75	<6.72e-05	1.93e-01	1.92e-01	1.93e-01	0.52	n/a	2.55e-05	n/a
S96T004003		Europium-154 by GEA	ug/mL	n/a	<7.64e-05	<1.08e-04	<8.75e-5	n/a	n/a	n/a	1.08e-04	n/a
S96T004003		Europium-155 by GEA	ug/mL	n/a	<7.88e-05	<6.03e-04	<6.09e-4	n/a	n/a	n/a	1.00e-03	n/a
S96T004003		Americium-241 by GEA	ug/mL	n/a	<1.72e-04	<1.25e-03	<1.24e-3	n/a	n/a	n/a	1.00e-03	n/a
S96T004003		Am-241 by Extraction	ug/mL	106.7	<5.44e-06	<5.32e-06	<5.46E-6	n/a	n/a	n/a	5.32e-06	1.00e+02
S96T004003		Cm-243/244 by Extraction	ug/mL	n/a	<5.44e-06	<5.32e-06	<5.46E-6	n/a	n/a	n/a	5.32e-06	1.00e+02

CORE NUMBER: 153
SEGMENT #: Blank

Table 1. Data Summary Table
AN-105

SEGMENT PORTION: Hydrostatic Head Fluid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003631	D	Silver-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	N/A	N/A
S961003631	D	Aluminum-ICP-Acid Dil.	ug/mL	96.00	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Arsenic-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Boron-ICP-Acid Dil.	ug/mL	99.20	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Barium-ICP-Acid Dil.	ug/mL	96.60	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Beryllium-ICP-Acid Dil.	ug/mL	101.8	<5.00e-03	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S961003631	D	Bismuth-ICP-Acid Dil.	ug/mL	100.2	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Calcium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Cadmium-ICP-Acid Dil.	ug/mL	99.20	<5.00e-03	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S961003631	D	Cerium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Cobalt-ICP-Acid Dil.	ug/mL	99.20	<2.00e-02	<2.00e-01	<2.00e-1	N/A	N/A	N/A	2.00e-01	N/A
S961003631	D	Chromium-ICP-Acid Dil.	ug/mL	100.2	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Copper-ICP-Acid Dil.	ug/mL	100.6	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Iron-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Potassium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-01	<5.00e-01	<5.00e0	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Lanthanum-ICP-Acid Dil.	ug/mL	97.60	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Lithium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Magnesium-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Molybdenum-ICP-Acid Dil.	ug/mL	99.60	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Sodium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Nickel-ICP-Acid Dil.	ug/mL	98.80	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Phosphorus-ICP-Acid Dil.	ug/mL	103.2	<2.00e-02	<2.00e-01	<2.00e-1	N/A	N/A	N/A	2.00e-01	N/A
S961003631	D	Lead-ICP-Acid Dil.	ug/mL	98.60	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Sulfur-ICP-Acid Dil.	ug/mL	99.60	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Antimony-ICP-Acid Dil.	ug/mL	94.80	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Selenium-ICP-Acid Dil.	ug/mL	96.60	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Silicon-ICP-Acid Dil.	ug/mL	95.80	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Strontium-ICP-Acid Dil.	ug/mL	95.20	<1.00e-01	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-1	N/A
S961003631	D	Titanium-ICP-Acid Dil.	ug/mL	96.20	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Thallium-ICP-Acid Dil.	ug/mL	96.60	<2.00e-01	<2.00e-01	<2.00e0	N/A	N/A	N/A	2.00e-01	N/A
S961003631	D	Uranium-ICP-Acid Dil.	ug/mL	92.80	<5.00e-01	<5.00e-01	<5.00e0	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Vanadium-ICP-Acid Dil.	ug/mL	97.80	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S961003631	D	Zinc-ICP-Acid Dil.	ug/mL	99.80	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Zirconium-ICP-Acid Dil.	ug/mL	102.7	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Fluoride-IC-Dioxex 4000/4500	ug/mL	103.0	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Chloride-IC-Dioxex 4000/4500	ug/mL	105.7	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Nitrite-IC - Dionex 4000/4500	ug/mL	102.0	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Bromide by Ion Chromatograph	ug/mL	102.5	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Nitrate by IC-Dionex 4000/4500	ug/mL	102.5	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-01	N/A
S961003631	D	Phosphate-IC-Dionex 4000/4500	ug/mL	99.82	<1.00e-01	<1.00e-01	<1.00e0	N/A	N/A	N/A	1.00e-01	N/A

13-Jan-1997 11:19:23
A-0002-3

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S96T003631		Sulfate by IC-Dionex 4000/4500	ug/mL	103.5	<1.36e-01	<1.51e+02	<1.51e+02	N/A	N/A	N/A	151.0	N/A
S96T003651		Oxalate by IC-Dionex 4000/4500	ug/mL	106.1	<1.05e-01	<1.17e+02	<1.17e+02	N/A	N/A	N/A	116.7	N/A

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

PHOTOGRAPHS

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

AN105 Core 152 Segment #1

6/17/96

AN103 Core 152 Segment #2

AN105 Core 152 Segment #6

AN-105 Core 152 Segment #6

6/14/90

AN-105 Core 152 Segment #7

6-27-96

6-27-96

AN-105 Core 152 Segment #8

AN-105 Core 152 Segment #9

6-27-96

WHC-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

AN-105 Core 152 Segment #10

6-27-96

AN-105 Core 152 Segment #11

6-27-96

WHC-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

AN-105 Core 152 Segment #12

6-27-96

WHC-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

AN-105 Core 152 Segment #13

6-27-96

WHO-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

7-1-96

AN-105 Core 152 Segment # 14

7-1-96

AN-105
Core 154 Segment #2-16

7-8-96

AN-105 Core 152 Segment #18

AN-105 Core 152 Segment #20

7/8/96

WHO-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

AN-105 Core 152 Segment #4632

7-1-96

8
100mm

AN-105 Core 153 Drill Blank

7-9-96

AN-105 Core 153 Segment #1

8-14-96

AN-105 Core 153 Segment #2

8/5/96

8-14-90

AN-105 Core 153 Segment #3

AN-105 Core 153 Segment #5

8/12/96

AN-105 Core 153 Segment #6

8-12-96

8-12-96

AN-105 Core 153 Segment #7

7-9-96

AN-105 Core 153 Segment #8

AN-105 Core 153 Segment #9

7/8/96

8/5/96

AN-105 Core 153 Segment #10

AN-105 Core 153 Segment #11 7/8/96

AN-105 Core 153 Segment #12

7/8/92

AN-105 Core 153 Segment #13

1996

8-7-96

AN-105 Core 153 Segment #14

8/5/96

AN-105 Core 153 Segment #75

AW-105 Core 153 Segment #17

8-14-96

AN-105 Core 153 Segment #19

8-14-96

7-9-96

AN-105 Core 153 Segment #21

7-9-96

AN-105 Core 153 Segment #22

HNF-SD-WM-OP-199, REV. 1

WHC-SD-WM-OP-199, REV. 1

SAMPLE HANDLING

HNH-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

Composite - Liquid Worksheet

Date: 7/24/96

Tank: PA105

Core 152

P.C. STEEN

Phone# 332-2495

Segment 1
Sample

ID: 590203715

Segment 2
Sample

ID: 590203716

Segment 3
Sample

ID: 590203717

Segment 4
Sample

ID: 590203718

Jar#: 10244

Jar/Vial Size: 250 mL

Initial Weight: 379.9 g

Final Weight: 377.6 g

Net Weight: 22.4 g

Jar#: 10080

Jar/Vial Size: 125 mL

Initial Weight: 173.0 g

Final Weight: 152.2 g

Net Weight: 20.8 g

Jar#: 10081

Jar/Vial Size: 125 mL

Initial Weight: 559.2 g

Final Weight: 537.8 g

Net Weight: 21.4 g

Jar#: 10077

Jar/Vial Size: 125 mL

Initial Weight: 557.0 g

Final Weight: 534.4 g

Net Weight: 22.6 g

Start Time:

End Time:

Tests:

Composite ✓

Md

1-8

Sample# 590204034

Jar#: 10634

Jar/Vial Size: 250 mL

Final Weight: 527.6 g

Initial Weight: 252.2 g

Net Weight: 304.4 g

Homogenization
Time (Min): N/A

(after final
subsample added)

Special Instructions: Build composite with 15 mLs from each sample. If first jar is depleted

use sample from jar listed in parenthesis.

* SAVE ALL REMAINING SAMPLE FOR MINUTIZATION PROGRAM

Composite - Liquid Worksheet

Date: 7/24/196

Tank: 4N105

Corr 152

P.C. STEEL

Phone# 322-249

Jar# 10072 (or 0245)
 Jar/Vial Size: 125 mL
 Initial Weight: 568.1 g
 Final Weight: 5462 g
 Net Weight: 21.9 g

Jar# 10082 (or 0246)
 Jar/Vial Size: 125 mL
 Initial Weight: 582.1 g
 Final Weight: 560.0 g
 Net Weight: 22.1 g

Jar# 10402
 Jar/Vial Size: 125 mL
 Initial Weight: 168.1 g
 Final Weight: 150.1 g
 Net Weight: 18.0 g

Jar# 10391
 Jar/Vial Size: 125 mL
 Initial Weight: 183.8 g
 Final Weight: 162.3 g
 Net Weight: 21.5 g

206

See Parent
 Jar# 10634
 on Page #1

Special Instructions:

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Age

2-1996

12-0176

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Age

2-1996

12-0176

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Age

2-1996

12-0176

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Age

2-1996

12-0176

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Age

2-1996

12-0176

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

WHC 2275 LAB ROOM 22

PACKS 108

Phone# 322-249

Segment 8 Sample

ID: 5847003722

Jar# 10391

Jar/Vial Size: 125 mL

Initial Weight: 183.8 g

Final Weight: 162.3 g

Net Weight: 21.5 g

Jar# 10402

Jar/Vial Size: 125 mL

Initial Weight: 168.1 g

Final Weight: 150.1 g

Net Weight: 18.0 g

Jar# 10082 (or 0246)

Jar/Vial Size: 125 mL

Initial Weight: 582.1 g

Final Weight: 560.0 g

Net Weight: 22.1 g

Jar# 10072 (or 0245)

Composite - Liquid Worksheet

Date: 7/24/96
Tank: AW105

P.C. 5/22/96
Phone# 372-24

Jar#: 10344
Jar/Vial Size: 250mL
Initial Weight: 529.3 g
Final Weight: 567.3 g
Net Weight: 22.0 g

Jar#: 10346
Jar/Vial Size: 250mL
Initial Weight: 545.9 g
Final Weight: 523.8 g
Net Weight: 22.1 g

Jar#: 10352 (or 10352)
Jar/Vial Size: 250mL
Initial Weight: 582.3 g
Final Weight: 560.2 g
Net Weight: 22.1 g

Jar#: 10348
Jar/Vial Size: 250mL
Initial Weight: 516.7 g
Final Weight: 493.2 g
Net Weight: 23.5 g

See Parent
Jar# 10334
on Page #1

Special Instructions:

Composite - Liquid Worksheet

Date: 7/24/96

Tank: AN105

Sample 13
15 mL ID: 547003727

Jar#: 10347
Jar/Vial Size: 250 mL
Initial Weight: 532.8 g
Final Weight: 509.3 g
Net Weight: 23.5 g

Sample 14
Sample ID: 547003728

Jar#: 10345
Jar/Vial Size: 250 mL
Initial Weight: 286.6 g
Final Weight: 264.7 g
Net Weight: 21.9 g

Sample ID:
Jar#: 9609
Jar/Vial Size: 1 mL
Initial Weight: g
Final Weight: g
Net Weight: g

Sample ID:
Jar#: 9609
Jar/Vial Size: 1 mL
Initial Weight: g
Final Weight: g
Net Weight: g

2000

See Parent
Jar# 10634
on Page #1

Special Instructions:

P.C. STEPA

Phone# 372-246

Aug 2 1996 12:01PM

WHC 2225 LAB ROOM 2F BACKSIDE

WHC-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

No. 3653

Composite Worksheet

Date: 8/2/96

Tank: AW105

Sample
ID: 594003738
Auger: 20gms
Core: 152
Segment: 14UH

Jar# 106822 (Bin 69)
Jar/Vial Size: 60.0 mL
Initial Weight: 174.1 g
Final Weight: 146.5 g
Net Weight: 27.6 g

Sample
ID: 594003799
Auger: 20gms
Core: 152
Segment: 14UH

Jar# 106855 (Bin 69)
Jar/Vial Size: 60.0 mL
Initial Weight: 113.7 g
Final Weight: 93.2 g
Net Weight: 20.5 g

Sample
ID: 594002803
Auger: 20gms
Core: 152
Segment: 16UH

Jar# 106816 (Bin 69)
Jar/Vial Size: 60.0 mL
Initial Weight: 192.9 g
Final Weight: 167.1 g
Net Weight: 25.8 g

Sample
ID: 5940038
Auger: 20gms
Core: 152
Segment: 16UH

Jar# 106877 (Bin 69)
Jar/Vial Size: 60.0 mL
Initial Weight: 188.5 g
Final Weight: 163.9 g
Net Weight: 20.6 g

P.C. Stan

Phone# 372-2069

WHC-SO-WM-DP-199, REV. 1

WHC-SO-WM-DP-199, REV. 1

Start Time: _____
End Time: _____

209

Tests:
COMBUST
MONITOR
FLAME

Jar# 10971
Jar/Vial Size: 125 mL
Final Weight: 281.4 g
Initial Weight: 120.3 g
Net Weight: 161.1 g (after final subsample added)

Homogenization Time (Min.): _____

(after final subsample added)

Special Instructions: Analyse all remaining sample in jar # 10685 (approx 8 grams after Compositing)

Build composite with 20 grams from each jar.

Composite Worksheet

Date: 8/2/96

Tank: BU105

P.C. STEW

Phone# 372-2495

Sample
ID: 5867003804
Auger/ 20gms
Core: 152
Segment: 18.1H

Sample
ID: 5867003811
Auger/ 20gms
Core: 152
Segment: 18.1H

Sample
ID: 5867003805
Auger/ 20gms
Core: 152
Segment: 20.1H

Sample
ID: 5867003812
Auger/ 20gms
Core: 152
Segment: 20.1H

Jar# 10688 (Bin 69)
Jar/Vial Size: 60 mL
Initial Weight: 122.2 g
Final Weight: 122.7 g
Net Weight: 20.5 g

Jar# 10690 (Bin 69)
Jar/Vial Size: 60 mL
Initial Weight: 149.1 g
Final Weight: 147.6 g
Net Weight: 21.5 g

Jar# 10691 (Bin 69)
Jar/Vial Size: 60 mL
Initial Weight: 145.4 g
Final Weight: 144.8 g
Net Weight: 20.6 g

Jar# 10692 (Bin 69)
Jar/Vial Size: 60 mL
Initial Weight: 148.6 g
Final Weight: 128.4 g
Net Weight: 21.2 g

WHC-SD-WM-DP-199, REV. 1

WHF-SD-WM-DP-199, REV. 1

See Parent
Jar# 10821
on Page #1

210

Special Instructions:

Composite % Calculation Sheet

Date: 8/17/96 Technician: A. J. LaPlant

- 1) Segment jar # 10682:

$$\frac{20.3 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.6 \%$$
- 2) Segment jar # 10685:

$$\frac{20.2 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.5 \%$$
- 3) Segment jar # 10686:

$$\frac{19.9 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.4 \%$$
- 4) Segment jar # 10687:

$$\frac{20.0 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.4 \%$$
- 5) Segment jar # 10688:

$$\frac{19.9 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.4 \%$$
- 6) Segment jar # 10690:

$$\frac{20.3 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.6 \%$$
- 7) Segment jar # 10691:

$$\frac{20.1 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.5 \%$$
- 8) Segment jar # 10692:

$$\frac{20.4 \text{ grams from segment jar to composite jar}}{161.1 \text{ Total grams in composite jar}} \times 100 = 12.7 \%$$
- 9) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$
- 10) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

Composite Worksheet

Date: 8/2/96

Tank: AN105

Sample
ID: SA0003806
Auger/
Core: 152
Segment: 22U4

Jar# 10694 (Bin 69)
Jar/Vial Size: 60 mL
Initial Weight: 157.2 g
Final Weight: 100.3 g
Net Weight: 50.9 g

Sample
ID: SA0003713
Auger/
Core: 152
Segment: 22L4

Jar# 10397 (Floor)
Jar/Vial Size: 125 mL
Initial Weight: 178.4 g
Final Weight: 134.6 g
Net Weight: 43.8 g

Sample
ID: _____
Auger/
Core: _____
Segment: _____

Jar# _____
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Sample
ID: _____
Auger/
Core: _____
Segment: _____

Jar# _____
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

P.C. SAEEN
Phone# 372-2495

Start Time: _____
End Time: _____

Tests:
COMBUST
HAZARDOUS
BASECATH

Jar# 10972
Jar/Vial Size: 125 mL
Final Weight: 213.1 g
Initial Weight: 140.2 g
Net Weight: 92.9 g

Homogenization
Time (Min.): _____

(after final subsample added)

Special Instructions: Build composite with entire sample from jar 10397 (Sample 3713). This jar should contain approx 50 grams, add the equal amount of 100 gm #10694 (approx 50 grams). Inhibit remaining sample in jar 10694 if possible.

Composite & Calculation Sheet

Date: 8/17/92 Technician: A. J. Pitt

- 1) Segment jar #
- 10694
- :

$$\frac{50.1 \text{ grams from segment jar to composite jar}}{92.9 \text{ Total grams in composite jar}} \times 100 = \underline{53.9} \%$$

- 2) Segment jar #
- 10357
- :

$$\frac{42.8 \text{ grams from segment jar to composite jar}}{92.9 \text{ Total grams in composite jar}} \times 100 = \underline{46.1} \%$$

- 3) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 4) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 5) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 6) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 7) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 8) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 9) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

- 10) Segment jar # _____:

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \underline{\hspace{2cm}} \%$$

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

Bulk Density Worksheet

Requestor: Spent

Date: 8/17/96

Tank: Ad65
Core: 152
Seg: Conf 14-20
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 16971
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Cone#: 71
Final Vol: 11.2 mL
Initial Weight: 7.5 g
Final Weight: 15.0 g
Net Weight: 17.5 g

Sample ID: 5927204207

Appearance/Narrative: Ball Density = 1.56

17.5 / 11.2 = 1.56

Tank: _____
Core: _____
Seg: _____
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: _____
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Cone#: _____
Final Vol: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

LABCORE Data Entry Template for Worklist#

9966

Analyst: CE Instrument: RA000 Book # NA

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C152 SEG # 1 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCK01		EXTRUS01	SOLID	<u>20</u>	<u>20.00</u>	N/A	
		2	INSTCK02		EXTRUS01	SOLID	<u>500</u>	<u>500.05</u>	N/A	
96000832	AN-105	3	SAMPLE	9961003481	0	DL10VOL1	SOLID	N/A	<u>190</u>	ml
96000832	AN-105	4	SAMPLE	9961003481	0	DL10WT01	SOLID	N/A	<u>296.8</u>	g
96000832	AN-105	5	SAMPLE	9961003481	0	EST. G/ML	SOLID	N/A	<u>1.35</u>	g/ml
96000832	AN-105	6	SAMPLE	9961003481	0	EXTRUS01	SOLID	N/A	<u>complete</u>	
96000832	AN-105	7	SAMPLE	9961003481	0	LL10WT01	SOLID	N/A	<u><5</u>	g
96000832	AN-105	8	SAMPLE	9961003481	0	NOTEROCK	SOLID	N/A	<u>N-1744</u>	
96000832	AN-105	9	SAMPLE	9961003481	0	SLOVOL01	SOLID	N/A	<u>NA</u>	ml
96000832	AN-105	10	SAMPLE	9961003481	0	SLOWT-01	SOLID	N/A	<u>37.6</u>	g
96000832	AN-105	11	SAMPLE	9961003481	0	APPEAR01	SOLID	N/A	<u>complete</u>	
96000832	AN-105	12	SAMPLE	9961003481	0	DRYVOL01	SOLID	N/A	<u>0</u>	ml

Final page for worklist #

9966

Analyst Signature CE Date 6-17-96

Analyst Signature CE Date 6-17-96

Reviewed by Steven B. Gandy 6-19-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Dimension Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9967

Analyst: EL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 10

Worklist Comment: AN-105 C152 SEG # 2 RISER 12A EXTRUSION

ORIG	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01				EXTRUD01	SOLID	<u>20</u>	<u>2000</u>	N/A	
		2	INSTRCK02				EXTRUD01	SOLID	<u>500</u>	<u>50000</u>	N/A	
96000832	AN-105	3	SAMPLE	9967003482	0		DLTUVOL1	SOLID	N/A	<u>305</u>		ml
96000832	AN-105	4	SAMPLE	9967003482	0		DLTUVOL1	SOLID	N/A	<u>406.3</u>		g
96000832	AN-105	5	SAMPLE	9967003482	0		EST. G/ML	SOLID	N/A	<u>139</u>		g/ml
96000832	AN-105	6	SAMPLE	9967003482	0		EXTRUD01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7	SAMPLE	9967003482	0		DLTUVOL1	SOLID	N/A	<u><5</u>		g
96000832	AN-105	8	SAMPLE	9967003482	0		NOTERBOOK	SOLID	N/A	<u>N-1344</u>		
96000832	AN-105	9	SAMPLE	9967003482	0		SLEWVOL1	SOLID	N/A	<u>NA</u>		ml
96000832	AN-105	10	SAMPLE	9967003482	0		SLEWVOL1	SOLID	N/A	<u>0</u>		g
96000832	AN-105	11	SAMPLE	9967003482	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12	SAMPLE	9967003482	0		ORIGVOL1	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

9967

Analyst Signature [Signature] Date 6-18-96 Analyst Signature [Signature] Date 6-18-96
Reviewed by [Signature] 6-17-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9968

Analyst: EC Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 62

Worklist Comment: AM-105 C152 SEG # 3 RISER 12A EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUS01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUS01	SOLID	<u>500</u>	<u>500.06</u>	<u>N/A</u>	
96000832	AM-105	3	SAMPLE	9961003483	0		DL10VOL1	SOLID	<u>N/A</u>	<u>230</u>		ml
96000832	AM-105	4	SAMPLE	9961003483	0		DL10WT01	SOLID	<u>N/A</u>	<u>406.2</u>		g
96000832	AM-105	5	SAMPLE	9961003483	0		EXT. G/M.	SOLID	<u>N/A</u>	<u>1.45</u>		g/ml
96000832	AM-105	6	SAMPLE	9961003483	0		EXTRUS01	SOLID	<u>N/A</u>	<u>complete</u>		
96000832	AM-105	7	SAMPLE	9961003483	0		DL10WT01	SOLID	<u>N/A</u>	<u>0</u>		g
96000832	AM-105	8	SAMPLE	9961003483	0		NOTESBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>		
96000832	AM-105	9	SAMPLE	9961003483	0		DL10VOL1	SOLID	<u>N/A</u>	<u>NA</u>		ml
96000832	AM-105	10	SAMPLE	9961003483	0		SLDWY-01	SOLID	<u>N/A</u>	<u>0</u>		g
96000832	AM-105	11	SAMPLE	9961003483	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
96000832	AM-105	12	SAMPLE	9961003483	0		DL10VOL1	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist #

9968

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9969

Analyst: EC Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 60

Worklist Comment: AN-105 C152 SEG # 4 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCKR01			EXTRUS01	SOLID	<u>20</u>	<u>2001</u>	<u>N/A</u>	
		2 INSTCKR02			EXTRUS01	SOLID	<u>500</u>	<u>500.0</u>	<u>N/A</u>	
96000832	AN-105	3 SAMPLE	9961003491	0	DLT0VOL1	SOLID	<u>N/A</u>	<u>300</u>		ml
96000832	AN-105	4 SAMPLE	9961003491	0	DLT0VOL1	SOLID	<u>N/A</u>	<u>400.3</u>		g
96000832	AN-105	5 SAMPLE	9961003491	0	EST. G/ML	SOLID	<u>N/A</u>	<u>1.43</u>		g/ml
96000832	AN-105	6 SAMPLE	9961003491	0	EXTRUS01	SOLID	<u>N/A</u>	<u>Complete</u>		
96000832	AN-105	7 SAMPLE	9961003491	0	LLT0VOL1	SOLID	<u>N/A</u>	<u>25</u>		g
96000832	AN-105	8 SAMPLE	9961003491	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>		
96000832	AN-105	9 SAMPLE	9961003491	0	SLD0VOL1	SOLID	<u>N/A</u>	<u>NA</u>		ml
96000832	AN-105	10 SAMPLE	9961003491	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
96000832	AN-105	11 SAMPLE	9961003491	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
96000832	AN-105	12 SAMPLE	9961003491	0	ORG0VOL1	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist #

9969

Analyst Signature: [Signature] Date: 6-18-96
Analyst Signature: [Signature] Date: 6-18-96
Reviewed by: [Signature] 6-19-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9970

Analyst: EL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod B0

Worklist Comment: AN-105 C152 SEG # 5 RISER 12A EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01				EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2	INSTRCK02				EXTRUD01	SOLID	<u>500</u>	<u>500.06</u>	N/A	
96000832	AN-105	3	SAMPLE	9961003492	0		DLTUVOL1	SOLID	N/A	<u>310</u>		ml
96000832	AN-105	4	SAMPLE	9961003492	0		DLTUVOL1	SOLID	N/A	<u>435.8</u>		g
96000832	AN-105	5	SAMPLE	9961003492	0		EST_G/ML	SOLID	N/A	<u>140</u>		g/ml
96000832	AN-105	6	SAMPLE	9961003492	0		EXTRUD01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7	SAMPLE	9961003492	0		DLTUVOL1	SOLID	N/A	<u>5</u>		g
96000832	AN-105	8	SAMPLE	9961003492	0		NOTED00K	SOLID	N/A	<u>N-1344</u>		
96000832	AN-105	9	SAMPLE	9961003492	0		SLOVOLD1	SOLID	N/A	<u>NA</u>		ml
96000832	AN-105	10	SAMPLE	9961003492	0		SLOWT-01	SOLID	N/A	<u>0</u>		g
96000832	AN-105	11	SAMPLE	9961003492	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12	SAMPLE	9961003492	0		ORGVOLD1	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

9970

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

9971

Analyst: EC Instrument: BA000 Book # 11A

Method: LO-160-103 Rev/Mod DO

Worklist Comment: AN-105 C152 SEG # 6 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FORM	DL	UNIT
		1	INSTCHK01		EXTRUSION	SOLID	<u>20</u>	<u>20.00</u>	<u>N/A</u>	
		2	INSTCHK02		EXTRUSION	SOLID	<u>500</u>	<u>500.06</u>	<u>N/A</u>	
96000832	AN-105	3	SAMPLE	9961003493	0	DLTQVOL1	SOLID	<u>N/A</u>	<u>290</u>	<u>ml</u>
96000832	AN-105	4	SAMPLE	9961003493	0	DLTQWT01	SOLID	<u>N/A</u>	<u>33.5</u>	<u>g</u>
96000832	AN-105	5	SAMPLE	9961003493	0	EST. G/Ml	SOLID	<u>N/A</u>	<u>1.48</u>	<u>g/ml</u>
96000832	AN-105	6	SAMPLE	9961003493	0	EXTRUSION	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	7	SAMPLE	9961003493	0	LLTQWT01	SOLID	<u>N/A</u>	<u>25</u>	<u>g</u>
96000832	AN-105	8	SAMPLE	9961003493	0	NOTESBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>	
96000832	AN-105	9	SAMPLE	9961003493	0	SLDWOL01	SOLID	<u>N/A</u>	<u>N/A</u>	<u>ml</u>
96000832	AN-105	10	SAMPLE	9961003493	0	SLDWL-01	SOLID	<u>N/A</u>	<u>0</u>	<u>g</u>
96000832	AN-105	11	SAMPLE	9961003493	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	12	SAMPLE	9961003493	0	DRZVOL01	SOLID	<u>N/A</u>	<u>0</u>	<u>ml</u>

Final page for worklist # 9971

Analyst Signature: EC Date: 6-24-96
 Analyst Signature: EC Date: 6-24-96
 Reviewed by: John B. Bony Date: 6-26-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9972

Analyst: EC Instrument: BA000 Book # NAMethod: LO-160-103 Rev/Mod 00

Worklist Comment: AN-105 C152 SEG # 7 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCKR1		EXTRUS01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCKR2		EXTRUS01	SOLID	<u>500</u>	<u>500.07</u>	<u>N/A</u>	
96000832	AN-105	3	SAMPLE	996T003494	0	DLTGVOL1	SOLID	<u>N/A</u>	<u>290</u>	ml
96000832	AN-105	4	SAMPLE	996T003494	0	DLTGVOL1	SOLID	<u>N/A</u>	<u>433.4</u>	g
96000832	AN-105	5	SAMPLE	996T003494	0	EST. G/M	SOLID	<u>N/A</u>	<u>1.49</u>	g/ml
96000832	AN-105	6	SAMPLE	996T003494	0	EXTRUS01	SOLID	<u>N/A</u>	<u>any/10</u>	
96000832	AN-105	7	SAMPLE	996T003494	0	LLTGVOL1	SOLID	<u>N/A</u>	<u>0</u>	g
96000832	AN-105	8	SAMPLE	996T003494	0	NOTESBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>	
96000832	AN-105	9	SAMPLE	996T003494	0	DLTGVOL1	SOLID	<u>N/A</u>	<u>NA</u>	ml
96000832	AN-105	10	SAMPLE	996T003494	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>	g
96000832	AN-105	11	SAMPLE	996T003494	0	APPEAR01	SOLID	<u>N/A</u>	<u>any/10</u>	
96000832	AN-105	12	SAMPLE	996T003494	0	DLTGVOL1	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist #

9972

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for Q17 (SPK & STD) may not reflect the actual unit. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

9973

Analyst: LL Instrument: BA000 Book # 11A

Method: LO-160-103 Rev/Mod 80

Worklist Comment: AN-105 C152 SEG # 8 RISER 12A EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01				EXTRUS01	SOLID	<u>20</u>	<u>30.01</u>	N/A	
		2	INSTRCK02				EXTRUS01	SOLID	<u>500</u>	<u>500.07</u>	N/A	
96000832	AN-105	3	SAMPLE	5961003495	0		OLIVOL1	SOLID	N/A	<u>300</u>		ml
96000832	AN-105	4	SAMPLE	5961003495	0		ULTRW101	SOLID	N/A	<u>421.3</u>		g
96000832	AN-105	5	SAMPLE	5961003495	0		EST./G/ML	SOLID	N/A	<u>1.43</u>		g/ml
96000832	AN-105	6	SAMPLE	5961003495	0		EXTRUS01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7	SAMPLE	5961003495	0		ULTRW101	SOLID	N/A	<u>0</u>		g
96000832	AN-105	8	SAMPLE	5961003495	0		ROTEROCK	SOLID	N/A	<u>N-1344</u>		
96000832	AN-105	9	SAMPLE	5961003495	0		SLEWOL01	SOLID	N/A	<u>NA</u>		ml
96000832	AN-105	10	SAMPLE	5961003495	0		SLEW-01	SOLID	N/A	<u>0</u>		g
96000832	AN-105	11	SAMPLE	5961003495	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12	SAMPLE	5961003495	0		ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

9973

Analyst Signature: [Signature] Date: 6-27-96
Analyst Signature: [Signature] Date: 6-27-96
Reviewed by: Arthur B. Penney 7-3-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10012

Analyst: EL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod B0

Worklist Comment: AN-105 C152 SEG # 9 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATELY	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUS01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02		EXTRUS01	SOLID	<u>500</u>	<u>500.07</u>	<u>N/A</u>	
96000832	AN-105	3	SAMPLE	S961003502	0	DL10VOL1	SOLID	<u>N/A</u>	<u>250</u>	ml
96000832	AN-105	4	SAMPLE	S961003502	0	DL10WTD1	SOLID	<u>N/A</u>	<u>378.7</u>	g
96000832	AN-105	5	SAMPLE	S961003502	0	EST. G/ML	SOLID	<u>N/A</u>	<u>1.51</u>	g/ml
96000832	AN-105	6	SAMPLE	S961003502	0	EXTRUS01	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	7	SAMPLE	S961003502	0	LL10WTD1	SOLID	<u>N/A</u>	<u><5</u>	g
96000832	AN-105	8	SAMPLE	S961003502	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>11-1344</u>	
96000832	AN-105	9	SAMPLE	S961003502	0	SL0VOL01	SOLID	<u>N/A</u>	<u>NA</u>	ml
96000832	AN-105	10	SAMPLE	S961003502	0	SL0WT-01	SOLID	<u>N/A</u>	<u>0</u>	g
96000832	AN-105	11	SAMPLE	S961003502	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	12	SAMPLE	S961003502	0	GR0VOL01	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist #

10012

Analyst Signature EL Date 6-27-96

Analyst Signature EL Date 6-27-96

Reviewed by Andrew B. Perry 7-3-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10014

Analyst: CL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod BO

Worklist Comment: AN-105 C152 SEG # 10 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUD01	SOLID	<u>20</u>	<u>20.00</u>	<u>N/A</u>	
		2	INSTCHK02		EXTRUD01	SOLID	<u>500</u>	<u>500.06</u>	<u>N/A</u>	
96000832	AN-105	3	SAMPLE	9967003503	0	DLTQV01	SOLID	<u>N/A</u>	<u>250</u>	ml
96000832	AN-105	4	SAMPLE	9967003503	0	DLTQV01	SOLID	<u>N/A</u>	<u>392.2</u>	g
96000832	AN-105	5	SAMPLE	9967003503	0	EST. G/M	SOLID	<u>N/A</u>	<u>1.56</u>	g/ml
96000832	AN-105	6	SAMPLE	9967003503	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	7	SAMPLE	9967003503	0	DLTQV01	SOLID	<u>N/A</u>	<u>25</u>	g
96000832	AN-105	8	SAMPLE	9967003503	0	NOTERBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>	
96000832	AN-105	9	SAMPLE	9967003503	0	SLOV01	SOLID	<u>N/A</u>	<u>NA</u>	ml
96000832	AN-105	10	SAMPLE	9967003503	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>	g
96000832	AN-105	11	SAMPLE	9967003503	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>	
96000832	AN-105	12	SAMPLE	9967003503	0	ORGV01	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist # 10014

Analyst Signature

Date

Analyst Signature

Date

Reviewed by: Kathleen B. Sanyal 7-3-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10015

Analyst: CC Instrument: BA000 Book # 169

Method: LO-160-103 Rev/Mod B0

Worklist Comment: AN-105 C152 SEG # 11 RISER 12A EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCD1				EXTRUD01	SOLID	<u>20</u>	<u>2000</u>	N/A	
		2	INSTONK02				EXTRUD01	SOLID	<u>500</u>	<u>50006</u>	N/A	
96000832	AN-105	3	SAMPLE	9967003504	0		OLIGVOL1	SOLID	N/A	<u>300</u>		ML
96000832	AN-105	4	SAMPLE	9967003504	0		OLIGVOL1	SOLID	N/A	<u>48.4</u>		g
96000832	AN-105	5	SAMPLE	9967003504	0		EST. G/ML	SOLID	N/A	<u>1.43</u>		g/ML
96000832	AN-105	6	SAMPLE	9967003504	0		EXTRUD01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7	SAMPLE	9967003504	0		LLIGVOL1	SOLID	N/A	<u>55</u>		g
96000832	AN-105	8	SAMPLE	9967003504	0		NOTEBOOK	SOLID	N/A	<u>11-1344</u>		
96000832	AN-105	9	SAMPLE	9967003504	0		SIGVOL01	SOLID	N/A	<u>N/A</u>		ML
96000832	AN-105	10	SAMPLE	9967003504	0		SIGVOL01	SOLID	N/A	<u>0</u>		g
96000832	AN-105	11	SAMPLE	9967003504	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12	SAMPLE	9967003504	0		ORGVOL01	SOLID	N/A	<u>0</u>		ML

Final page for worklist # 10015

Analyst Signature

Date

Analyst Signature

Date

Reviewed by Arthur B. Bondy 7-3-96

Data Entry Comments

LABCORE Data Entry Template for Worklist#

10110

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C152 SEG # 18 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01		EXTRUD01	SOLID	<u>20</u>	<u>20.51</u>	N/A	
		2	INSTRCK02		EXTRUD01	SOLID	<u>500</u>	<u>500.07</u>	N/A	
96000832	AN-105	3	SAMPLE	9961003505	0	DI1QV01	SOLID	N/A		ml
96000832	AN-105	4	SAMPLE	9961003505	0	DI1QV01	SOLID	N/A		g
96000832	AN-105	5	SAMPLE	9961003505	0	EST.G/ML	SOLID	N/A		g/ml
96000832	AN-105	6	SAMPLE	9961003505	0	EXTRUD01	SOLID	N/A	<u>Complete</u>	
96000832	AN-105	7	SAMPLE	9961003505	0	DI1QV01	SOLID	N/A	<u><5 above 7/09/96</u>	
96000832	AN-105	8	SAMPLE	9961003505	0	NOTEBOOK	SOLID	N/A	<u>N-1704</u>	
96000832	AN-105	9	SAMPLE	9961003505	0	SLDV01	SOLID	N/A	<u>N/A</u>	ml
96000832	AN-105	10	SAMPLE	9961003505	0	SLDV-01	SOLID	N/A	<u>983.5</u>	g
96000832	AN-105	11	SAMPLE	9961003505	0	APPEAR01	SOLID	N/A	<u>Complete</u>	
96000832	AN-105	12	SAMPLE	9961003505	0	GRV01	SOLID	N/A	<u>22</u>	ml

Final page for worklist #

10110

Susan Bee 7/09/96
Analyst Signature Date

Susan Bee 7/09/96
Analyst Signature Date

Bethand Jiffy
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10181

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-O

Worklist Comment: AN-105 C152 SEG # 14 RISER 12A EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20</u>	<u>20.00</u>	N/A	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500</u>	<u>500.08</u>	N/A	
96000832	AN-105	3	SAMPLE	9961003512	0		DLTUVOL1	SOLID	N/A	<u>70</u>		ml
96000832	AN-105	4	SAMPLE	9961003512	0		DLTUVOL1	SOLID	N/A	<u>132.9</u>		g
96000832	AN-105	5	SAMPLE	9961003512	0		EST. G/MIL	SOLID	N/A	<u>1.48</u>		g/oz.
96000832	AN-105	6	SAMPLE	9961003512	0		EXTRUD01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7	SAMPLE	9961003512	0		DLTUVOL1	SOLID	N/A	<u>6.5</u>		g
96000832	AN-105	8	SAMPLE	9961003512	0		NOTEBOOK	SOLID	N/A	<u>N-1394</u>		
96000832	AN-105	9	SAMPLE	9961003512	0		SLOVOLD1	SOLID	N/A	<u>NA</u>		ml
96000832	AN-105	10	SAMPLE	9961003512	0		SLEWT-01	SOLID	N/A	<u>23.4</u>		g
96000832	AN-105	11	SAMPLE	9961003512	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12	SAMPLE	9961003512	0		ORGVOLD1	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

10181

Analyst Signature Susan B. Bue Date 7/09/96

Analyst Signature Susan B. Bue Date 7/09/96

Reviewed by Arthur B. Bue 7/17/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10182

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-O

Worklist Comment: AN-105 C152 SEG # 16 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUD01	SOLID	<u>32</u>	<u>20.00</u>	N/A	
		2	INSTCHK02		EXTRUD01	SOLID	<u>520</u>	<u>520.08</u>	N/A	
96000832	AN-105	3	SAMPLE	9961003513	0	DL10VOL1	SOLID	N/A	<u>0</u>	ml
96000832	AN-105	4	SAMPLE	9961003513	0	DL10W101	SOLID	N/A	<u>0</u>	g
96000832	AN-105	5	SAMPLE	9961003513	0	EST. G/M	SOLID	N/A	<u>NA</u>	g/ml
96000832	AN-105	6	SAMPLE	9961003513	0	EXTRUD01	SOLID	N/A	<u>complete</u>	
96000832	AN-105	7	SAMPLE	9961003513	0	DL10W101	SOLID	N/A	<u>5</u>	g
96000832	AN-105	8	SAMPLE	9961003513	0	ROTATION	SOLID	N/A	<u>11.1344</u>	
96000832	AN-105	9	SAMPLE	9961003513	0	SLEWVOL1	SOLID	N/A	<u>NA</u>	ml
96000832	AN-105	10	SAMPLE	9961003513	0	SLENT-01	SOLID	N/A	<u>376.8</u>	g
96000832	AN-105	11	SAMPLE	9961003513	0	APPEAR01	SOLID	N/A	<u>complete</u>	
96000832	AN-105	12	SAMPLE	9961003513	0	ORGVOL01	SOLID	N/A	<u>0</u>	ml

Final page for worklist #

10182

Analyst Signature Susan Bee Date 7/09/96

Analyst Signature Susan Bee Date 7/09/96

Richard B. Johnson Jr
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10183

Analyst: Blair Instrument: BA000 Book # _____Method: LO-160-103 Rev/Mod B-V

Worklist Comment: AN-105 C152 SEG # 20 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTENK01		EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2	INSTENK02		EXTRUD01	SOLID	<u>500</u>	<u>500.07</u>	N/A	
96000832	AN-105	3	SAMPLE	9967003514	0	DLTUVOL1	SOLID	N/A	<u>0</u>	ML
96000832	AN-105	4	SAMPLE	9967003514	0	DLTQWTO1	SOLID	N/A	<u>0</u>	g
96000832	AN-105	5	SAMPLE	9967003514	0	EST, G/ML	SOLID	N/A	<u>NA</u>	g/ml
96000832	AN-105	6	SAMPLE	9967003514	0	EXTRUD01	SOLID	N/A	<u>Complete</u>	
96000832	AN-105	7	SAMPLE	9967003514	0	LITISU101	SOLID	N/A	<u>7/10/96</u>	g
96000832	AN-105	8	SAMPLE	9967003514	0	NOTERBOOK	SOLID	N/A	<u>12.1344</u>	
96000832	AN-105	9	SAMPLE	9967003514	0	SLEVOLO1	SOLID	N/A	<u>NA</u>	ML
96000832	AN-105	10	SAMPLE	9967003514	0	SLEW1-01	SOLID	N/A	<u>596.9</u>	g
96000832	AN-105	11	SAMPLE	9967003514	0	APPEAR01	SOLID	N/A	<u>Complete</u>	
96000832	AN-105	12	SAMPLE	9967003514	0	ORGVOLO1	SOLID	N/A	<u>0</u>	ML

Final page for worklist #

10183

Analyst Signature: Blair Date: 7/10/96Analyst Signature: Blair Date: 7/10/96Bethann D. Griffin II
7-10-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10388

Analyst: EE Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod BO

Worklist Comment: AN-105 C152 SEG # 12 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01		EXTRUD01	SOLID	20	20.00	N/A	
		2	INSTRCK02		EXTRUD01	SOLID	500	500.06	N/A	
96000832	AN-105	3	SAMPLE	996T003588	G	DLT0VOL1	SOLID	N/A	250	ml
96000832	AN-105	4	SAMPLE	996T003588	G	DLT0WT01	SOLID	N/A	380.0	g
96000832	AN-105	5	SAMPLE	996T003588	G	EST, G/ML	SOLID	N/A	1.52	g/ml
96000832	AN-105	6	SAMPLE	996T003588	G	EXTRUD01	SOLID	N/A	complete	
96000832	AN-105	7	SAMPLE	996T003588	G	DLT0WT01	SOLID	N/A	25	g
96000832	AN-105	8	SAMPLE	996T003588	G	NOTEBOOK	SOLID	N/A	N-D/H	
96000832	AN-105	9	SAMPLE	996T003588	G	SLOV0L01	SOLID	N/A	N/A	ml
96000832	AN-105	10	SAMPLE	996T003588	G	SLOWT-C1	SOLID	N/A	0	g
96000832	AN-105	11	SAMPLE	996T003588	G	APPEAR01	SOLID	N/A	complete	
96000832	AN-105	12	SAMPLE	996T003588	G	OROV0L01	SOLID	N/A	0	ml

Final page for worklist #

10388

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10389

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod 6-0

Worklist Comment: AN-105 C152 SEG # 13 RISER 12A EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHKE01			EXTRUD01	SOLID	<u>2.0</u>	<u>20.00</u>	N/A	
		2 INSTCHKE2			EXTRUD01	SOLID	<u>200</u>	<u>200.06</u>	N/A	
96000832	AN-105	3 SAMPLE	9967003589	0	DL1RVOL1	SOLID	N/A	<u>250</u>		ml
96000832	AN-105	4 SAMPLE	9967003589	0	DL1FWT01	SOLID	N/A	<u>383.8</u>		g
96000832	AN-105	5 SAMPLE	9967003589	0	EST. G/ML	SOLID	N/A	<u>154</u>		g/ml
96000832	AN-105	6 SAMPLE	9967003589	0	EXTRUD01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	7 SAMPLE	9967003589	0	DL1FWT01	SOLID	N/A	<u>15</u>		g
96000832	AN-105	8 SAMPLE	9967003589	0	NOTEBOOK	SOLID	N/A	<u>N-1744</u>		
96000832	AN-105	9 SAMPLE	9967003589	0	SLOWVOL01	SOLID	N/A	<u>NA</u>		ml
96000832	AN-105	10 SAMPLE	9967003589	0	SLOWT-01	SOLID	N/A	<u>0</u>		g
96000832	AN-105	11 SAMPLE	9967003589	0	APPEAR01	SOLID	N/A	<u>complete</u>		
96000832	AN-105	12 SAMPLE	9967003589	0	OROVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist # 10389

Supra Bae 7/10/96
Analyst Signature Date

Supra Bae 7/10/96
Analyst Signature Date

Reviewed by John B. Bae 7/17/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10390

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C152 SEG # 22 RISER 12A EXTRUSION

ORIG	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUS01	SOLID	20	20.00	N/A	
		2	INSTCHK02		EXTRUS01	SOLID	225	225.09	N/A	
96000832	AN-105	3	SAMPLE	9961003590	0	DLTUVOL1	SOLID	N/A	0	nd.
96000832	AN-105	4	SAMPLE	9961003590	0	DLTUVOL1	SOLID	N/A	0	S
96000832	AN-105	5	SAMPLE	9961003590	0	EST. G/ML	SOLID	N/A	NA	g/ml
96000832	AN-105	6	SAMPLE	9961003590	0	EXTRUS01	SOLID	N/A	complete	
96000832	AN-105	7	SAMPLE	9961003590	0	DLTUVOL1	SOLID	N/A	0	S
96000832	AN-105	8	SAMPLE	9961003590	0	NOTEBOOK	SOLID	N/A	N-1344	
96000832	AN-105	9	SAMPLE	9961003590	0	SLOVOL01	SOLID	N/A	N/A	nd.
96000832	AN-105	10	SAMPLE	9961003590	0	SLOWT-01	SOLID	N/A	246.8	S
96000832	AN-105	11	SAMPLE	9961003590	0	APPEAR01	SOLID	N/A	complete	
96000832	AN-105	12	SAMPLE	9961003590	0	ORISVOL1	SOLID	N/A	0	nd.

Final page for worklist #

10390

Analyst Signature Steven B. Bandy Date 7/10/96

Analyst Signature Steven B. Bandy Date 7/10/96

Reviewed by Steven B. Bandy 7/10/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10611

Analyst: SLH Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod B-2

Worklist Comment: AN-105 CORE 153 SEG.22 EXTRUSION RH

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCKR01		EXTRUSION	SOLID	<u>50</u>	<u>89.01</u>	<u>N/A</u>	
		2	INSTCKR02		EXTRUSION	SOLID	<u>50</u>	<u>89.03</u>	<u>N/A</u>	
96000945	AN-105	3	SAMPLE	9961003630	0	DL10VOL1	SOLID	<u>N/A</u>	<u>50</u>	ML
96000945	AN-105	4	SAMPLE	9961003630	0	DL10WTD1	SOLID	<u>N/A</u>	<u>50</u>	g
96000945	AN-105	5	SAMPLE	9961003630	0	EST. G/ML	SOLID	<u>N/A</u>	<u>2.4</u>	g/86
96000945	AN-105	6	SAMPLE	9961003630	0	EXTRUSION	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	9961003630	0	LL10WTD1	SOLID	<u>N/A</u>	<u><10</u>	g
96000945	AN-105	8	SAMPLE	9961003630	0	NOTERBOOK	SOLID	<u>N/A</u>	<u>2.1344</u>	
96000945	AN-105	9	SAMPLE	9961003630	0	SL10VOL1	SOLID	<u>N/A</u>	<u>1/4</u>	ML
96000945	AN-105	10	SAMPLE	9961003630	0	SL10WT-01	SOLID	<u>N/A</u>	<u>257.3</u>	g
96000945	AN-105	11	SAMPLE	9961003630	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	9961003630	0	DR10VOL1	SOLID	<u>N/A</u>	<u>50</u>	ML

Final page for worklist #

10611

Analyst Signature SLH Date 7-15-96

Analyst Signature SLH Date 7-15-96

Reviewed by Arthur B. Bernaty 7-17-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10613

Analyst: SKB Instrument: RA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C153 Seg 9 Extrusion GL

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCK01		EXTRUS01	SOLID	<u>50</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCK02		EXTRUS01	SOLID	<u>500</u>	<u>200.07</u>	<u>N/A</u>	
96000945	AN-105	3	SAMPLE	5961003637	0	DL10VOL1	SOLID	<u>N/A</u>	<u>30</u>	ML
96000945	AN-105	4	SAMPLE	5961003637	0	DL10WT01	SOLID	<u>N/A</u>	<u>438.9</u>	g/100g
96000945	AN-105	5	SAMPLE	5961003637	0	EST G/ML	SOLID	<u>N/A</u>	<u>46.94</u>	g/ml
96000945	AN-105	6	SAMPLE	5961003637	0	EXTRUS01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	5961003637	0	DL10WT01	SOLID	<u>N/A</u>	<u>45</u>	g/100g
96000945	AN-105	8	SAMPLE	5961003637	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>11-1244</u>	
96000945	AN-105	9	SAMPLE	5961003637	0	SLDVOL01	SOLID	<u>N/A</u>	<u>N/A</u>	ML
96000945	AN-105	10	SAMPLE	5961003637	0	SLDWT-01	SOLID	<u>N/A</u>	<u>N/A</u>	g
96000945	AN-105	11	SAMPLE	5961003637	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	5961003637	0	DRGVOL01	SOLID	<u>N/A</u>	<u>Complete</u>	ML

Final page for worklist #

10613

Analyst Signature: Susan Bar Date: 7/6/96

Analyst Signature: Susan Bar Date: 7/6/96

Bernard Giffen II
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Set Number, R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10614

Analyst: SLB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C153 Seg 11 Extrusion GL

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01		EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTRCK02		EXTRUD01	SOLID	<u>500</u>	<u>500.07</u>	<u>N/A</u>	
96000945	AN-105	3	SAMPLE	9961003638	0	DL10VOL1	SOLID	<u>N/A</u>	<u>270</u>	ml
96000945	AN-105	4	SAMPLE	9961003638	0	DL10VOL1	SOLID	<u>N/A</u>	<u>420.0</u>	g
96000945	AN-105	5	SAMPLE	9961003638	0	EST. G/ML	SOLID	<u>N/A</u>	<u>1.44</u>	g/ml
96000945	AN-105	6	SAMPLE	9961003638	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	9961003638	0	DL10VOL1	SOLID	<u>N/A</u>	<u>LS</u>	g
96000945	AN-105	8	SAMPLE	9961003638	0	NOTEROCK	SOLID	<u>N/A</u>	<u>N-1324</u>	
96000945	AN-105	9	SAMPLE	9961003638	0	DL10VOL1	SOLID	<u>N/A</u>	<u>N/A</u>	ml
96000945	AN-105	10	SAMPLE	9961003638	0	SIDMT-01	SOLID	<u>N/A</u>	<u>2A</u>	g
96000945	AN-105	11	SAMPLE	9961003638	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	9961003638	0	DRYVOL01	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist # 10614

Susan Ben 7/09/96
Analyst Signature Date

Susan Ben 7/09/96
Analyst Signature Date

Bertina Duffin II
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10615

Analyst: SLB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod 6-0

Worklist Comment: AN-105 C153 SEG.12 EXTRUSION RH

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCHG1				EXTRUS01	SOLID	<u>SD</u>	<u>SD.01</u>	N/A	
		2	INSTRCHG2				EXTRUS01	SOLID	<u>SD</u>	<u>SD.01</u>	N/A	
96000945	AN-105	3	SAMPLE	9961003639	0		DL10V01	SOLID	N/A	<u>SD</u>		ml
96000945	AN-105	4	SAMPLE	9961003639	0		DL10V01	SOLID	N/A	<u>401.2</u>		g
96000945	AN-105	5	SAMPLE	9961003639	0		EST. G/M	SOLID	N/A	<u>1.40</u>		g/ml
96000945	AN-105	6	SAMPLE	9961003639	0		EXTRUS01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	7	SAMPLE	9961003639	0		DL10V01	SOLID	N/A	<u>LS</u>		g
96000945	AN-105	8	SAMPLE	9961003639	0		NOTED00K	SOLID	N/A	<u>4.1344</u>		
96000945	AN-105	9	SAMPLE	9961003639	0		SLDV01	SOLID	N/A	<u>NA</u>		ml
96000945	AN-105	10	SAMPLE	9961003639	0		SLDV01	SOLID	N/A	<u>NA</u>		g
96000945	AN-105	11	SAMPLE	9961003639	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	12	SAMPLE	9961003639	0		ORGV01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

10615

Analyst Signature SLB Date 7/17/96

Analyst Signature SLB Date 7/17/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10652

Analyst: SLB Instrument: RA000 Book # _____Method: LO-160-103 Rev/Mod 12-0

Worklist Comment: AN-105 C153 Seg FB Extrusion gll

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01		EXTRUS01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTRCK02		EXTRUS01	SOLID	<u>500</u>	<u>500.03</u>	<u>N/A</u>	
96000945	AN-105	3	SAMPLE	9961003663	0	DL10VOL1	SOLID	<u>N/A</u>	<u>250</u>	ml
96000945	AN-105	4	SAMPLE	9961003663	0	DL10WT01	SOLID	<u>N/A</u>	<u>272.7</u>	g
96000945	AN-105	5	SAMPLE	9961003663	0	EST. G/ML	SOLID	<u>N/A</u>	<u>1.11</u>	g/ml
96000945	AN-105	6	SAMPLE	9961003663	0	EXTRUS01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	9961003663	0	LL10WT01	SOLID	<u>N/A</u>	<u>0</u>	g
96000945	AN-105	8	SAMPLE	9961003663	0	NOTES00K	SOLID	<u>N/A</u>	<u>11-1244</u>	
96000945	AN-105	9	SAMPLE	9961003663	0	SLD1VOL01	SOLID	<u>N/A</u>	<u>NA</u>	ml
96000945	AN-105	10	SAMPLE	9961003663	0	SLDWT-01	SOLID	<u>N/A</u>	<u>NA</u>	g
96000945	AN-105	11	SAMPLE	9961003663	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	9961003663	0	DR01VOL01	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist #

10652

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10653

Analyst: SKB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C153 Seg 8 Exnsion gil

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCK01		EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2	INSTCK02		EXTRUD01	SOLID	<u>500</u>	<u>500.03</u>	N/A	
96000945	AN-105	3	SAMPLE	9967003664	0	OLIGVOL1	SOLID	N/A	<u>290</u>	ng
96000945	AN-105	4	SAMPLE	9967003664	0	OLIGVOL1	SOLID	N/A	<u>426.7</u>	g
96000945	AN-105	5	SAMPLE	9967003664	0	EST. G/ML	SOLID	N/A	<u>1.47</u>	g/ml
96000945	AN-105	6	SAMPLE	9967003664	0	EXTRUD01	SOLID	N/A	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	9967003664	0	OLIGVOL1	SOLID	N/A	<u>5</u>	g
96000945	AN-105	8	SAMPLE	9967003664	0	NOTERBOOK	SOLID	N/A	<u>6-1344</u>	
96000945	AN-105	9	SAMPLE	9967003664	0	SLOVOL01	SOLID	N/A	<u>N/A</u>	ng
96000945	AN-105	10	SAMPLE	9967003664	0	SLENT-01	SOLID	N/A	<u>N/A</u>	g
96000945	AN-105	11	SAMPLE	9967003664	0	APPEAR01	SOLID	N/A	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	9967003664	0	ORGVOL01	SOLID	N/A	<u>0</u>	ng

Final page for worklist #

10653

Susan Bar 7/09/96
Analyst Signature Date

Susan Bar 7/10/96
Analyst Signature Date

Bertand J. Buffin
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units DL = Detection Limit, S = Worklist Shot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

10654

Analyst: SAB Instrument: BA000 Book # _____Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C153 Sep 13 Extrusion gll

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTWKT01		EXTRUD01	SOLID	<u>200</u>	<u>80.01</u>	N/A	
		2	INSTWKT02		EXTRUD01	SOLID	<u>500</u>	<u>500.03</u>	N/A	
96000945	AN-105	3	SAMPLE	5961003665	0	DLTQV01	SOLID	<u>N/A</u>	<u>300</u>	ML
96000945	AN-105	4	SAMPLE	5961003665	0	DLTQW01	SOLID	<u>N/A</u>	<u>431.4</u>	g
96000945	AN-105	5	SAMPLE	5961003665	0	EST.G/ML	SOLID	<u>N/A</u>	<u>1.43</u>	g/mL
96000945	AN-105	6	SAMPLE	5961003665	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	7	SAMPLE	5961003665	0	DLTQW01	SOLID	<u>N/A</u>	<u>15</u>	g
96000945	AN-105	8	SAMPLE	5961003665	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>	
96000945	AN-105	9	SAMPLE	5961003665	0	SLOV01	SOLID	<u>N/A</u>	<u>NA</u>	ML
96000945	AN-105	10	SAMPLE	5961003665	0	SLOWT-01	SOLID	<u>N/A</u>	<u>NA</u>	g
96000945	AN-105	11	SAMPLE	5961003665	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
96000945	AN-105	12	SAMPLE	5961003665	0	DEGVOLO1	SOLID	<u>N/A</u>	<u>0</u>	ML

Final page for worklist #

10654

Susan Bee 7/12/96
 Analyst Signature Date

Susan Bee 7/12/96
 Analyst Signature Date

Entirely off
7-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

10655

Analyst: SAB Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod B-0

Worklist Comment: AN-105 C153 Seg 21 Extrusion gil

GROUP	PROJECT	S TYPE	SAMPLER#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCH01			EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2 INSTCH02			EXTRUD01	SOLID	<u>500</u>	<u>500.03</u>	N/A	
96000945	AN-105	3 SAMPLE	596T003666	0	DLT0V01	SOLID	N/A	<u>180</u>		ml
96000945	AN-105	4 SAMPLE	596T003666	0	DLT0W01	SOLID	N/A	<u>173</u>		g
96000945	AN-105	5 SAMPLE	596T003666	0	EST. G/M	SOLID	N/A	<u>1.44</u>		g/ml
96000945	AN-105	6 SAMPLE	596T003666	0	EXTRUD01	SOLID	N/A	<u>Complete</u>		
96000945	AN-105	7 SAMPLE	596T003666	0	DLT0W01	SOLID	N/A	<u><5</u>		g
96000945	AN-105	8 SAMPLE	596T003666	0	NOTEBOOK	SOLID	N/A	<u>0.1344</u>		
96000945	AN-105	9 SAMPLE	596T003666	0	SLOVEX01	SOLID	N/A	<u>N/A</u>		ml
96000945	AN-105	10 SAMPLE	596T003666	0	SLODWT01	SOLID	N/A	<u>202.5</u>		g
96000945	AN-105	11 SAMPLE	596T003666	0	APPEAR01	SOLID	N/A	<u>Complete</u>		
96000945	AN-105	12 SAMPLE	596T003666	0	ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

10655

Analyst Signature Susan Bui Date 7/12/96
Bertina Duff
7-12-96

Analyst Signature Susan Bui Date 7/12/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit. S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11399

Analyst: EL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 60

Worklist Comment: AN-105 C153 SEG # 2 RISER 7B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FORMD	DL	UNIT
		1	INSTRCHK01				EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTRCHK02				EXTRUD01	SOLID	<u>500</u>	<u>500.06</u>	<u>N/A</u>	
96000945	AN-105	3	SAMPLE	S96T004154	0		DL10VOL1	SOLID	<u>N/A</u>	<u>320</u>		ML
96000945	AN-105	4	SAMPLE	S96T004154	0		DL10WT01	SOLID	<u>N/A</u>	<u>422.8</u>		g
96000945	AN-105	5	SAMPLE	S96T004154	0		EST.G/ML	SOLID	<u>N/A</u>	<u>1.32</u>		g/ML
96000945	AN-105	6	SAMPLE	S96T004154	0		EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	7	SAMPLE	S96T004154	0		DL10WT01	SOLID	<u>N/A</u>	<u>55</u>		g
96000945	AN-105	8	SAMPLE	S96T004154	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>N-1244</u>		
96000945	AN-105	9	SAMPLE	S96T004154	0		SLDVOL01	SOLID	<u>N/A</u>	<u>NA</u>		ML
96000945	AN-105	10	SAMPLE	S96T004154	0		SLOWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
96000945	AN-105	11	SAMPLE	S96T004154	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	12	SAMPLE	S96T004154	0		DROVD01	SOLID	<u>N/A</u>	<u>0</u>		ML

Final page for worklist #

11399

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit. S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

242

LABCORE Data Entry Template for Worklist#

11400

Analyst: EL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod B0

Worklist Comment: AN-105 C153 SEG # 10 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUS01	SOLID	<u>20</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUS01	SOLID	<u>500</u>	<u>500.01</u>	<u>N/A</u>	
96000945	AN-105	3 SAMPLE	9961004155	0	ULTRAVOL 1	SOLID	<u>N/A</u>	<u>290</u>		ML
96000945	AN-105	4 SAMPLE	9961004155	0	ULTRAVOL 1	SOLID	<u>N/A</u>	<u>386.3</u>		g
96000945	AN-105	5 SAMPLE	9961004155	0	EST. G/ML	SOLID	<u>N/A</u>	<u>1.33</u>		g/ML
96000945	AN-105	6 SAMPLE	9961004155	0	EXTRUS01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	7 SAMPLE	9961004155	0	ULTRAVOL 1	SOLID	<u>N/A</u>	<u>< 5</u>		g
96000945	AN-105	8 SAMPLE	9961004155	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-1744</u>		
96000945	AN-105	9 SAMPLE	9961004155	0	ULTRAVOL 1	SOLID	<u>N/A</u>	<u>NA</u>		ML
96000945	AN-105	10 SAMPLE	9961004155	0	SLOWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
96000945	AN-105	11 SAMPLE	9961004155	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	12 SAMPLE	9961004155	0	ULTRAVOL 1	SOLID	<u>N/A</u>	<u>0</u>		ML

Final page for worklist #

11400

Analyst Signature EL Date 8-6-96
Bernard J. Giffen II.
8-7-96

Analyst Signature EL Date 8-6-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11401

Analyst: EC Instrument: BA000 Book # NAMethod: LO-160-103 Rev/Mod 62

Worklist Comment: AN-105 C153 SEG # 15 RISER 7B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTRCK01				EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2	INSTRCK02				EXTRUD01	SOLID	<u>500</u>	<u>500.06</u>	N/A	
96000945	AN-105	3	SAMPLE	596T004156	0		DI10VOL1	SOLID	N/A	<u>0</u>		ml
96000945	AN-105	4	SAMPLE	596T004156	0		DI10WTD1	SOLID	N/A	<u>0</u>		g
96000945	AN-105	5	SAMPLE	596T004156	0		EST.G/NL	SOLID	N/A	<u>0</u>		g/ml
96000945	AN-105	6	SAMPLE	596T004156	0		EXTRUD01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	7	SAMPLE	596T004156	0		DI10WTD1	SOLID	N/A	<u><5</u>		g
96000945	AN-105	8	SAMPLE	596T004156	0		NOTESBOOK	SOLID	N/A	<u>11/344</u>		
96000945	AN-105	9	SAMPLE	596T004156	0		DI10VOL01	SOLID	N/A	<u>NA</u>		ml
96000945	AN-105	10	SAMPLE	596T004156	0		SLOWT-01	SOLID	N/A	<u>379.4</u>		g
96000945	AN-105	11	SAMPLE	596T004156	0		APPEAR01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	12	SAMPLE	596T004156	0		DI10VOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

11401

EC 8-6-96
 Analyst Signature Date
Edward Biffin II
8-6-96

EC 8-6-96
 Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

11693

Analyst: CL Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod D0

Worklist Comment: AN-105 C153 SEG # 14 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHRD1			EXTRUS01	SOLID	20	20.01	N/A	
		2 INSTCHRD2			EXTRUS01	SOLID	500	500.08	N/A	
96000945	AN-105	3 SAMPLE	596T004324	0	DLTGVOL1	SOLID	N/A	100		ml
96000945	AN-105	4 SAMPLE	596T004324	0	DLTGVOL1	SOLID	N/A	130.1		g
96000945	AN-105	5 SAMPLE	596T004324	0	EST. G/ML	SOLID	N/A	1.30		g/ml
96000945	AN-105	6 SAMPLE	596T004324	0	EXTRUS01	SOLID	N/A	complete		
96000945	AN-105	7 SAMPLE	596T004324	0	DLTGVOL1	SOLID	N/A	<5		g
96000945	AN-105	8 SAMPLE	596T004324	0	NOTEBOOK	SOLID	N/A	11-1344		
96000945	AN-105	9 SAMPLE	596T004324	0	SLOVOL01	SOLID	N/A	NA		ml
96000945	AN-105	10 SAMPLE	596T004324	0	SLOVOL01	SOLID	N/A	248.6		g
96000945	AN-105	11 SAMPLE	596T004324	0	APPEAR01	SOLID	N/A	complete		
96000945	AN-105	12 SAMPLE	596T004324	0	GREVOL01	SOLID	N/A	0		ml

Final page for worklist #

11693

Analyst Signature CL Date 8-12-96

Analyst Signature CL Date 8-12-96

Reviewed by Stephen B. Gentry 8-12-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

11702

Analyst: EL Instrument: BA000 Book # N/AMethod: LO-160-103 Rev/Mod 50

Worklist Comment: AN-105 C153 SEG # 7 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500</u>	<u>500.07</u>	<u>N/A</u>	
96000945	AN-105	3 SAMPLE	S961004323	0	DL1GVOL1	SOLID	<u>N/A</u>	<u>295</u>		ml
96000945	AN-105	4 SAMPLE	S961004323	0	DL1GVOL1	SOLID	<u>N/A</u>	<u>402.6</u>		g
96000945	AN-105	5 SAMPLE	S961004323	0	EST. G/ML	SOLID	<u>N/A</u>	<u>136</u>		g/ml
96000945	AN-105	6 SAMPLE	S961004323	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	7 SAMPLE	S961004323	0	LL1GVOL1	SOLID	<u>N/A</u>	<u>LS</u>		g
96000945	AN-105	8 SAMPLE	S961004323	0	NOTEROCK	SOLID	<u>N/A</u>	<u>N/A</u>		
96000945	AN-105	9 SAMPLE	S961004323	0	SLOWVOL1	SOLID	<u>N/A</u>	<u>N/A</u>		ml
96000945	AN-105	10 SAMPLE	S961004323	0	SLOWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
96000945	AN-105	11 SAMPLE	S961004323	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	12 SAMPLE	S961004323	0	URCVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist #

11702

Analyst Signature EL Date 8-14-96Analyst Signature EL Date 8-14-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11780

Analyst: EC Instrument: BA000 Book # 1/A

Method: LO-160-103 Rev/Mod 00

Worklist Comment: AN-105 C153 SEG # 5 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20</u>	<u>20.00</u>	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	<u>5.10</u>	<u>5.0007</u>	N/A	
96000945	AN-105	3 SAMPLE	5961004326	0	DL1QV01	SOLID	N/A	<u>3.0</u>		ml
96000945	AN-105	4 SAMPLE	5961004326	0	DL1QV101	SOLID	N/A	<u>40.0</u>		g
96000945	AN-105	5 SAMPLE	5961004326	0	EST. G/ML	SOLID	N/A	<u>1.3</u>		g/ml
96000945	AN-105	6 SAMPLE	5961004326	0	EXTRUD01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	7 SAMPLE	5961004326	0	LL1QV101	SOLID	N/A	<u>0</u>		g
96000945	AN-105	8 SAMPLE	5961004326	0	NOTEBOOK	SOLID	N/A	<u>N-1344</u>		
96000945	AN-105	9 SAMPLE	5961004326	0	SLEW01	SOLID	N/A	<u>NA</u>		ml
96000945	AN-105	10 SAMPLE	5961004326	0	SLEW-01	SOLID	N/A	<u>0</u>		g
96000945	AN-105	11 SAMPLE	5961004326	0	APPEAR01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	12 SAMPLE	5961004326	0	ORGV01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

11780

Analyst Signature EC Date 8-14-96
Betha
8-14-96

Analyst Signature EC Date 8-14-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11781

Analyst: EC Instrument: BA000 Book # N/AMethod: LO-160-103 Rev/Mod 00

Worklist Comment: AN-105 C153 SEG # 6 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTEAD01			EXTRUD01	SOLID	<u>20</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTEAD02			EXTRUD01	SOLID	<u>500</u>	<u>500.07</u>	<u>N/A</u>	
96000945	AN-105	3 SAMPLE	596T004327	0	DLTQV01	SOLID	<u>N/A</u>	<u>300</u>		ml
96000945	AN-105	4 SAMPLE	596T004327	0	DLTQV01	SOLID	<u>N/A</u>	<u>400.7</u>		g
96000945	AN-105	5 SAMPLE	596T004327	0	EST G/ML	SOLID	<u>N/A</u>	<u>1.26</u>		g/ml
96000945	AN-105	6 SAMPLE	596T004327	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	7 SAMPLE	596T004327	0	DLTQV01	SOLID	<u>N/A</u>	<u>5</u>		g
96000945	AN-105	8 SAMPLE	596T004327	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-1344</u>		
96000945	AN-105	9 SAMPLE	596T004327	0	SLOV01	SOLID	<u>N/A</u>	<u>N/A</u>		ml
96000945	AN-105	10 SAMPLE	596T004327	0	SLOV01	SOLID	<u>N/A</u>	<u>0</u>		g
96000945	AN-105	11 SAMPLE	596T004327	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
96000945	AN-105	12 SAMPLE	596T004327	0	ORDV01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist #

11781

Analyst Signature EC Date 8-14-96Analyst Signature EC Date 8-14-96Bertrand Duffin II
8-14-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11782

Analyst: CE Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 30

Worklist Comment: AN-105 C153 SEG # 19 RISER 7B EXTRUSION

SLOT#	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20</u>	<u>20.01</u>	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500</u>	<u>500.11</u>	N/A	
96000945	AN-105	3 SAMPLE	5961004328	0	DL1QV01	SOLID	N/A	<u>30</u>		ml
96000945	AN-105	4 SAMPLE	5961004328	0	DL1QW101	SOLID	N/A	<u>0</u>		g
96000945	AN-105	5 SAMPLE	5961004328	0	EST.G/ML	SOLID	N/A	<u>0</u>		g/ML
96000945	AN-105	6 SAMPLE	5961004328	0	EXTRUD01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	7 SAMPLE	5961004328	0	DL1QW101	SOLID	N/A	<u>210</u>		g
96000945	AN-105	8 SAMPLE	5961004328	0	NOTEBOOK	SOLID	N/A	<u>12-1344</u>		
96000945	AN-105	9 SAMPLE	5961004328	0	SLDVOL01	SOLID	N/A	<u>NA</u>		ml
96000945	AN-105	10 SAMPLE	5961004328	0	SLDWT-01	SOLID	N/A	<u>244.1</u>		g
96000945	AN-105	11 SAMPLE	5961004328	0	APPEAR01	SOLID	N/A	<u>complete</u>		
96000945	AN-105	12 SAMPLE	5961004328	0	ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist #

11782

Analyst Signature: CE Date: 8-16-96
Bethand J. H. II
8-16-96

Analyst Signature: CE Date: 8-16-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11932

Analyst: CE Instrument: BA000 Book # N/AMethod: LO-160-103 Rev/Mod 30

Worklist Comment: AN-105 C153 SEG 1 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUD01	SOLID	20	20.01	N/A	
		2	INSTCHK02		EXTRUD01	SOLID	500	500.11	N/A	
96000945	AN-105	3	SAMPLE	5961004384	0	DL10VOL1	SOLID	N/A	165	ML
96000945	AN-105	4	SAMPLE	5961004384	0	DL10VOL1	SOLID	N/A	242.2	g
96000945	AN-105	5	SAMPLE	5961004384	0	EST. C/M	SOLID	N/A	1.41	g/ML
96000945	AN-105	6	SAMPLE	5961004384	0	EXTRUD01	SOLID	N/A	complete	
96000945	AN-105	7	SAMPLE	5961004384	0	DL10VOL1	SOLID	N/A	25	g
96000945	AN-105	8	SAMPLE	5961004384	0	NOTEBOOK	SOLID	N/A	11-1344	
96000945	AN-105	9	SAMPLE	5961004384	0	SLOVOL01	SOLID	N/A	N/A	ML
96000945	AN-105	10	SAMPLE	5961004384	0	SLENT-G1	SOLID	N/A	0	g
96000945	AN-105	11	SAMPLE	5961004384	0	APPEAR01	SOLID	N/A	complete	
96000945	AN-105	12	SAMPLE	5961004384	0	ORGVOL01	SOLID	N/A	0	ML

Final page for worklist #

11932

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11934

Analyst: EC

Instrument: BA000

Book # NAMethod: LO-160-103 Rev/Mod 00

Worklist Comment: AN-105 C153 SEG # 3 RISER 7B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	S A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCH01		EXTRUS01	SOLID	20	20.01	N/A	
		2	INSTCH02		EXTRUS01	SOLID	500	500.11	N/A	
96000945	AN-105	3	SAMPLE	996T004385	0	DLTUVOL1	SOLID	N/A	250	ml
96000945	AN-105	4	SAMPLE	996T004385	0	DLTUVOL1	SOLID	N/A	367.0	g
96000945	AN-105	5	SAMPLE	996T004385	0	EST. G/ML	SOLID	N/A	146	g/ml
96000945	AN-105	6	SAMPLE	996T004385	0	EXTRUS01	SOLID	N/A	complete	
96000945	AN-105	7	SAMPLE	996T004385	0	DLTUVOL1	SOLID	N/A	25	g
96000945	AN-105	8	SAMPLE	996T004385	0	NOTER00K	SOLID	N/A	N-1344	
96000945	AN-105	9	SAMPLE	996T004385	0	SLOVCL01	SOLID	N/A	NA	ml
96000945	AN-105	10	SAMPLE	996T004385	0	SLOWT-01	SOLID	N/A	0	g
96000945	AN-105	11	SAMPLE	996T004385	0	APPEAR01	SOLID	N/A	complete	
96000945	AN-105	12	SAMPLE	996T004385	0	OROVOL01	SOLID	N/A	0	ml

Final page for worklist #

11934

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11936

Analyst: EC Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod 80

Worklist Comment: AN-105 C153 SEG # 17 RISER 7B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTWKT01				EXTRUD01	SOLID	20	20.01	N/A	
		2	INSTWKT02				EXTRUD01	SOLID	500	500.11	N/A	
96000945	AN-105	3	SAMPLE	996T004386	0		DLTUVOL1	SOLID	N/A	0		ml
96000945	AN-105	4	SAMPLE	996T004386	0		DLTUVOL1	SOLID	N/A	0		g
96000945	AN-105	5	SAMPLE	996T004386	0		EST. 5/ML	SOLID	N/A	0		g/ml
96000945	AN-105	6	SAMPLE	996T004386	0		EXTRUD01	SOLID	N/A	complete		
96000945	AN-105	7	SAMPLE	996T004386	0		DLTUVOL1	SOLID	N/A	5		g
96000945	AN-105	8	SAMPLE	996T004386	0		NOTEBOOK	SOLID	N/A	N-1344		
96000945	AN-105	9	SAMPLE	996T004386	0		SLOVOL01	SOLID	N/A	N/A		ml
96000945	AN-105	10	SAMPLE	996T004386	0		SLWT-01	SOLID	N/A	421.1		g
96000945	AN-105	11	SAMPLE	996T004386	0		APPEAR01	SOLID	N/A	complete		
96000945	AN-105	12	SAMPLE	996T004386	0		ORGVOL01	SOLID	N/A	0		ml

Final page for worklist #

11936

Analyst Signature: EC Date: 8-16-96
Beckman
8-16-96

Analyst Signature: EC Date: 8-16-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

SAMPLE PREPARATIONS

HNH-SD-WM-OP-199, REV. 1

WHC-SD-WM-OP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

Verp Version 2.1 05/13/95

24/96 15.35

HNF-SD-WM-DP-199, REV. 1

LABCORE Data Entry Template for Worklist#

Analyst: BCD Instrument: FUS01 Book # _____Method: LA-549-141 Rev/Mod F.O.

Worklist Comment: AN-105 C152 SEG1/140/14L/16U FUSION01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	N A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>.250</u>	N/A	g/L
96000832	AN-105	2 SAMPLE	996T003827	0 F	FUSION01	SOLID	N/A	<u>2.0368</u>	<u>2.0368</u>	g/L
			<u>.5072g</u>							
96000832	AN-105	3 SAMPLE	996T003827	0 F	DOSE-02	SOLID	N/A	<u>.65</u>		mrads/hour
			<u>.4970g</u>							
96000832	AN-105	4 DUP	996T003827	0 F	FUSION01	SOLID	<u>2.0368</u>	<u>1.988</u>	N/A	g/L
			<u>.5072g</u>							
96000832	AN-105	5 DUP	996T003827	0 F	DOSE-02	SOLID	<u>.65</u>	<u>.65</u>	N/A	mrads/hour
			<u>.5185g</u>							
96000832	AN-105	6 SAMPLE	996T003893	0 F	FUSION01	SOLID	N/A	<u>2.074</u>		g/L
			<u>.5185g</u>							
96000832	AN-105	7 SAMPLE	996T003893	0 F	DOSE-02	SOLID	N/A	<u>.60</u>		mrads/hour
			<u>.5099g</u>							
96000832	AN-105	8 DUP	996T003893	0 F	FUSION01	SOLID	<u>2.074</u>	<u>2.0396</u>	N/A	g/L
			<u>.5031g</u>							
96000832	AN-105	9 DUP	996T003893	0 F	DOSE-02	SOLID	<u>.60</u>	<u>.60</u>	N/A	mrads/hour
			<u>.5113g</u>							
96000832	AN-105	10 SAMPLE	996T003894	0 F	FUSION01	SOLID	N/A	<u>2.0124</u>		g/L
			<u>.5031g</u>							
96000832	AN-105	11 SAMPLE	996T003894	0 F	DOSE-02	SOLID	N/A	<u>.65</u>		mrads/hour
			<u>.5113g</u>							
96000832	AN-105	12 DUP	996T003894	0 F	FUSION01	SOLID	<u>2.0124</u>	<u>2.0532</u>	N/A	g/L
			<u>.5113g</u>							
96000832	AN-105	13 DUP	996T003894	0 F	DOSE-02	SOLID	<u>.65</u>	<u>.65</u>	N/A	mrads/hour
			<u>.5013g</u>							
96000832	AN-105	14 SAMPLE	996T003895	0 F	FUSION01	SOLID	N/A	<u>2.0052</u>		g/L
			<u>.5113g</u>							
96000832	AN-105	15 SAMPLE	996T003895	0 F	DOSE-02	SOLID	N/A	<u>.50</u>		mrads/hour
			<u>.5103g</u>							
96000832	AN-105	16 DUP	996T003895	0 F	FUSION01	SOLID	<u>2.0052</u>	<u>2.0052</u>	N/A	g/L
			<u>.5103g</u>							
96000832	AN-105	17 DUP	996T003895	0 F	DOSE-02	SOLID	<u>.50</u>	<u>.50</u>	N/A	mrads/hour

Data Entry Comments:

.596T003897 → .3827
.3731 → .3192
.3816 → .3894
.3792 → .3445

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

255

PK Miller
 8/12/96

LABCORE Data Entry Template for Worklist#

11117

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
-------	---------	--------	---------	-----	------	--------	--------	-------	----	------

Final page for worklist #

11117

D. J. D. D. 8-5-96
Analyst Signature Date

R. J. D. 8-9-96
Analyst Signature Date

Reviewed by R. K. Fuller
8/12/96

Data Entry Comments:

Units shown for QC (SPE & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11146

Analyst: DeD Instrument: FUS01 Book # _____Method: LA-549-141 Rev/Mod F-0

Worklist Comment: AN105 C152 SEG 16L/18U/18L/200 FUSION01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLANK PREP			FUSION01	SOLID	1	250	N/A	g/L
96000832	AN-105	2 SAMPLE	596T003889	0 F	FUSION01	SOLID	N/A	2.0212		g/L
			5053g →		2.50L					
96000832	AN-105	3 SAMPLE	596T003889	0 F	DOSE-02	SOLID	N/A	60		mrads/hour
96000832	AN-105	4 DUP	596T003889	0 F	FUSION01	SOLID	20212	2.046	N/A	g/L
			5115g →		2.50L					
96000832	AN-105	5 DUP	596T003889	0 F	DOSE-02	SOLID	60	60	N/A	mrads/hour
96000832	AN-105	6 SAMPLE	596T003890	0 F	FUSION01	SOLID	N/A	1.9912		g/L
			4978g →		2.50L					
96000832	AN-105	7 SAMPLE	596T003890	0 F	DOSE-02	SOLID	N/A	50		mrads/hour
96000832	AN-105	8 DUP	596T003890	0 F	FUSION01	SOLID	1.9912	1.98	N/A	g/L
			4950g →		2.50L					
96000832	AN-105	9 DUP	596T003890	0 F	DOSE-02	SOLID	50	50	N/A	mrads/hour
96000832	AN-105	10 SAMPLE	596T003896	0 F	FUSION01	SOLID	N/A	1.9844		g/L
			4961g →		2.50L					
96000832	AN-105	11 SAMPLE	596T003896	0 F	DOSE-02	SOLID	N/A	50		mrads/hour
96000832	AN-105	12 DUP	596T003896	0 F	FUSION01	SOLID	1.9844	2.0416	N/A	g/L
			5104g →		2.50L					
96000832	AN-105	13 DUP	596T003896	0 F	DOSE-02	SOLID	50	50	N/A	mrads/hour
96000832	AN-105	14 SAMPLE	596T003897	0 F	FUSION01	SOLID	N/A	1.934		g/L
			4835g →		2.50L					
96000832	AN-105	15 SAMPLE	596T003897	0 F	DOSE-02	SOLID	N/A	50		mrads/hour
96000832	AN-105	16 DUP	596T003897	0 F	FUSION01	SOLID	1.934	1.98	N/A	g/L
			4950g →		2.50L					
96000832	AN-105	17 DUP	596T003897	0 F	DOSE-02	SOLID	50	50	N/A	mrads/hour

596T003898 → 3899
 5995 → 3899
 5994 → 3896
 5996 → 3899

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11146

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
-------	---------	--------	---------	-----	------	--------	--------	-------	----	------

Final page for worklist #

11146

Daniel D. D... 8-5-96
Analyst Signature Date

E. J. ... 8-9-96
Analyst Signature Date

Data Entry Comments:

Reviewed by R K Fells
8/12/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11147

Analyst: DPR Instrument: FUS01 Book # Method: LA-549-141 Rev/Mod F-0Worklist Comment: AN105 C152 SEG20L/220 FUSION01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FORMD	DL	UNIT
		1 BLANK PREP			FUSION01	SOLID	1	250	N/A	g/L
96000832	AN-105	2 SAMPLE	960008391	0 F	FUSION01	SOLID	N/A	2,0084		g/L
96000832	AN-105	3 SAMPLE	960008391	0 F	DOSE-02	SOLID	N/A	65		mgd/hour
96000832	AN-105	4 DUP	960008391	0 F	FUSION01	SOLID	2,0084	1,9936	N/A	g/L
96000832	AN-105	5 DUP	960008391	0 F	DOSE-02	SOLID	65	65	N/A	mgd/hour
96000832	AN-105	6 SAMPLE	960008392	0 F	FUSION01	SOLID	N/A	1,9628		g/L
96000832	AN-105	7 SAMPLE	960008392	0 F	DOSE-02	SOLID	N/A	80		mgd/hour
96000832	AN-105	8 DUP	960008392	0 F	FUSION01	SOLID	1,9628	1,93	N/A	g/L
96000832	AN-105	9 DUP	960008392	0 F	DOSE-02	SOLID	80	92	N/A	mgd/hour

Final page for worklist #

11147

David P. Bromley 8/5/96
Analyst Signature Date

[Signature] 8-9-96
Analyst Signature Date

2967008997 → 3841
5798 → 3822

Data Entry Comments:

Reviewed by *PK Feller*
8/12/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

12107

Analyst: ADP Instrument: FUS01 Book # _____

Method: LA-549-141 Rev/Mod F-6

Worklist Comment: AN-105 Fusion Digests C153 Seg 14 LR, 15 LH, UH lad

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	1	250	N/A	g/L
96000945	AN-105	2 SAMPLE	596T004314 0 F		FUSION01	SOLID	N/A	2.151		g/L
			.5377 → 250L							
96000945	AN-105	3 SAMPLE	596T004314 0 F		DOSE-02	SOLID	N/A	60		mgad/hour
96000945	AN-105	4 DUP	596T004314 0 F		FUSION01	SOLID	2.151	2.139	N/A	g/L
			.5347 → 250L							
96000945	AN-105	5 DUP	596T004314 0 F		DOSE-02	SOLID	60	60	N/A	mgad/hour
96000945	AN-105	6 SAMPLE	596T004317 0 F		FUSION01	SOLID	N/A	2.054		g/L
			.5135 → 250L							
96000945	AN-105	7 SAMPLE	596T004317 0 F		DOSE-02	SOLID	N/A	60		mgad/hour
96000945	AN-105	8 DUP	596T004317 0 F		FUSION01	SOLID	2.054	2.018	N/A	g/L
			.5046 → 250L							
96000945	AN-105	9 DUP	596T004317 0 F		DOSE-02	SOLID	60	60	N/A	mgad/hour
96000945	AN-105	10 SAMPLE	596T004337 0 F		FUSION01	SOLID	N/A	2.018		g/L
			.5112 → 250L							
96000945	AN-105	11 SAMPLE	596T004337 0 F		DOSE-02	SOLID	N/A	60		mgad/hour
96000945	AN-105	12 DUP	596T004337 0 F		FUSION01	SOLID	2.018	2.020	N/A	g/L
			.5054 → 250L							
96000945	AN-105	13 DUP	596T004337 0 F		DOSE-02	SOLID	60	60	N/A	mgad/hour

Final page for worklist #

12107

ADP 8-26-96
Analyst Signature Date

ADP 8-26-96
Analyst Signature Date

TR Ok 8-26-96

Data Entry Comments
UP - Bill Bishop
10ml con H2O3, <190 dpg/dl

Received by R. Keller
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

12108

Analyst: ADD Instrument: FUS01 Book # _____Method: LA-549-141 Rev/Mod F-0

Worklist Comment: AN-105 Fusion Digest C153 Seg 17 UH, LR, 19 LH lad

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FORMO	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	1	250	N/A	g/L
96000945	AN-105	2 SAMPLE	5961004441 0 F		FUSION01	SOLID	N/A	2091		g/L
			• 5227 → 250							
96000945	AN-105	3 SAMPLE	5961004441 0 F		DOSE-02	SOLID	N/A	60		mrad/hour
96000945	AN-105	4 DUP	5961004441 0 F		FUSION01	SOLID	2091	2087	N/A	g/L
			• 5217 → 250							
96000945	AN-105	5 DUP	5961004441 0 F		DOSE-02	SOLID	60	60	N/A	mrad/hour
96000945	AN-105	6 SAMPLE	5961004442 0 F		FUSION01	SOLID	N/A	2035		g/L
			• 5087 → 250							
96000945	AN-105	7 SAMPLE	5961004442 0 F		DOSE-02	SOLID	N/A	60		mrad/hour
96000945	AN-105	8 DUP	5961004442 0 F		FUSION01	SOLID	2025	2019	N/A	g/L
			• 5122 → 250							
96000945	AN-105	9 DUP	5961004442 0 F		DOSE-02	SOLID	60	60	N/A	mrad/hour
96000945	AN-105	10 SAMPLE	5961004443 0 F		FUSION01	SOLID	N/A	2028		g/L
			• 5070 → 250							
96000945	AN-105	11 SAMPLE	5961004443 0 F		DOSE-02	SOLID	N/A	60		mrad/hour
96000945	AN-105	12 DUP	5961004443 0 F		FUSION01	SOLID	2028	2018	N/A	g/L
			• 5119 → 250							
96000945	AN-105	13 DUP	5961004443 0 F		DOSE-02	SOLID	60	60	N/A	mrad/hour

Final page for worklist #

12108

Anthony Purno 8-26-96
Analyst Signature Date

TR Olson 8-26-96
Analyst Signature Date

Data Entry Comments:

10 ml con H₂O₂, 190 g/L
10 ml con H₂O₂, 190 g/L

Received by RK Fuller 8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

12109

Analyst: ADP Instrument: FUS01 Book # Method: LA-549-141 Rev/Mod F-0Worklist Comment: AN-105 Fusion Digest C153 Seg 21 LH, 22 LH, UH lad

GROUP	PROJECT	S TYPE	SAMPLEN	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>250</u>	<u>N/A</u>	<u>g/L</u>
96000945	AN-105	2 SAMPLE	596T003984	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2057</u>		<u>g/L</u>
			<u>.5092 → 250L</u>							
96000945	AN-105	3 SAMPLE	596T003984	0 F	DOSE-02	SOLID	<u>N/A</u>	<u>60</u>		<u>mrad/hour</u>
96000945	AN-105	4 DUP	596T003984	0 F	FUSION01	SOLID	<u>2037</u>	<u>2013</u>	<u>N/A</u>	<u>g/L</u>
			<u>.5107 → 250L</u>							
96000945	AN-105	5 DUP	596T003984	0 F	DOSE-02	SOLID	<u>60</u>	<u>60</u>	<u>N/A</u>	<u>mrad/hour</u>
96000945	AN-105	6 SAMPLE	596T003985	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2036</u>		<u>g/L</u>
			<u>.5091 → 250L</u>							
96000945	AN-105	7 SAMPLE	596T003985	0 F	DOSE-02	SOLID	<u>N/A</u>	<u>60</u>		<u>mrad/hour</u>
96000945	AN-105	8 DUP	596T003985	0 F	FUSION01	SOLID	<u>2036</u>	<u>2036</u>	<u>N/A</u>	<u>g/L</u>
			<u>.5090 → 250L</u>							
96000945	AN-105	9 DUP	596T003985	0 F	DOSE-02	SOLID	<u>60</u>	<u>60</u>	<u>N/A</u>	<u>mrad/hour</u>
96000945	AN-105	10 SAMPLE	596T003986	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2053</u>		<u>g/L</u>
			<u>.5133 → 250L</u>							
96000945	AN-105	11 SAMPLE	596T003986	0 F	DOSE-02	SOLID	<u>N/A</u>	<u>60</u>		<u>mrad/hour</u>
96000945	AN-105	12 DUP	596T003986	0 F	FUSION01	SOLID	<u>2033</u>	<u>2019</u>	<u>N/A</u>	<u>g/L</u>
			<u>.5042 → 250L</u>							
96000945	AN-105	13 DUP	596T003986	0 F	DOSE-02	SOLID	<u>60</u>	<u>60</u>	<u>N/A</u>	<u>mrad/hour</u>

Final page for worklist #

12109

Analyst Signature: [Signature] Date: 8-26-96Analyst Signature: [Signature] Date: 8-26-96
T.R. Olson 08-27-96

Data Entry Comments:

W.P. Bishop
40 ml con H₂O₂, 5196 DALTONReviewed by RK Fuller
8/28/96Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist# 12383

Analyst: AKB Instrument: FDS01 Book # _____

Method: LA-549-141 Rev/Mod F-0

Worklist Comment: AN-105 Fusion Digest Seg 14-20 Core Comp lad and Seg 22 Core Comp Rkt 8/24/96

GROUP	PROJECT	Z TYPE	SAMPLE#	U A	TEST	MATRIX	ACTUAL	FOURD	DL	UNIT
		1	BLK-PRP		FUSION01	SOLID	<u>1</u>	<u>250</u>	N/A	g/L
96000832	AN-105	2	SAMPLE	596T004256	Q F	FUSION01	SOLID	N/A	<u>2.0312</u>	g/L
				<u>.9078</u>						
96000832	AN-105	3	SAMPLE	596T004256	Q F	DOSE-02	SOLID	N/A	<u>75</u>	mrad/hour
96000832	AN-105	4	DUP	596T004256	Q F	FUSION01	SOLID	<u>2.0312</u>	<u>2.0536</u>	N/A
				<u>.91343</u>						
96000832	AN-105	5	DUP	596T004256	Q F	DOSE-02	SOLID	<u>75</u>	<u>75</u>	N/A
96000832	AN-105	6	SAMPLE	596T004261	Q F	FUSION01	SOLID	N/A	<u>2.1464</u>	g/L
				<u>.93669</u>						
96000832	AN-105	7	SAMPLE	596T004261	Q F	DOSE-02	SOLID	N/A	<u>50</u>	mrad/hour
96000832	AN-105	8	DUP	596T004261	Q F	FUSION01	SOLID	<u>2.1464</u>	<u>1.9848</u>	N/A
				<u>.94629</u>						
96000832	AN-105	9	DUP	596T004261	Q F	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A

Final page for worklist # 12383

Analyst Signature

Date

Analyst Signature

Date

596T004254 → 596T004256 14-20 Comp
4259 → 4261 22 Comp

Reviewed by RK Miller
8/24/96

Data Entry Comments:

Some con H₂O₂ < 10% solids

HPT- Scott Corbally

Submitted BLM

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

LABCORE Data Entry Template for Worklist#

11150

Analyst: DGP Instrument: H2001 Book # _____

Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN105 C152 SEG10/14U H20DIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2001G01	SOLID	1.0100	N/A		g/L
96000832	AN-105	2 SAMPLE	5961003907 0 W 5.032g →		H2001G01	SOLID	N/A	5.032		g/L
96000832	AN-105	3 SAMPLE	5961003907 0 W		DOSE-02	SOLID	N/A	60		mead/hour
96000832	AN-105	4 DUP	5961003907 0 W 4.999g →		H2001G01	SOLID	5.032	4.999	N/A	g/L
96000832	AN-105	5 DUP	5961003907 0 W		DOSE-02	SOLID	60	60	N/A	mead/hour
96000832	AN-105	6 SAMPLE	5961003915 0 W 5.144g →		H2001G01	SOLID	N/A	5.144		g/L
96000832	AN-105	7 SAMPLE	5961003915 0 W		DOSE-02	SOLID	N/A	60		mead/hour
96000832	AN-105	8 DUP	5961003915 0 W 5.165g →		H2001G01	SOLID	5.144	5.165	N/A	g/L
96000832	AN-105	9 DUP	5961003915 0 W		DOSE-02	SOLID	60	55	N/A	mead/hour

Final page for worklist #

11150

DGP 8-5-96
Analyst Signature Date

R Jones 8-10-96
Analyst Signature Date

Rae Ann Green 8-9-96

5961003915 → 3975
3816 → 5907

Reviewed by RK Fuller
8/12/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11151

Analyst: DGD Instrument: H2001 Book # _____Method: LA-504-101 Rev/Mod RG E-0

Worklist Comment: AN105 C152 SEG14L/16U H2ODIG01 SKB

GROW#	PROJECT	S TYPE	SAMPLE#	N A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP				H2001G01	SOLID	<u>1</u>	<u>0.100</u>	N/A g/L
96000832	AN-105	2 SAMPLE	S96T003829 <u>5044g</u> →	0 W		H2001G01	SOLID	N/A	<u>5044</u>	g/L
96000832	AN-105	3 SAMPLE	S96T003829 <u>4963g</u> →	0 W	DOSE-02	SOLID	N/A	<u>40</u>	<u>4963</u>	mgad/hour
96000832	AN-105	4 DUP	S96T003829 <u>4963g</u> →	0 W		H2001G01	SOLID	<u>5044</u>	<u>4963</u>	N/A g/L
96000832	AN-105	5 DUP	S96T003829 <u>4963g</u> →	0 W	DOSE-02	SOLID	<u>40</u>	<u>40</u>	N/A	mgad/hour
96000832	AN-105	6 SAMPLE	S96T003908 <u>4926g</u> →	0 W		H2001G01	SOLID	N/A	<u>4926</u>	g/L
96000832	AN-105	7 SAMPLE	S96T003908 <u>4926g</u> →	0 W	DOSE-02	SOLID	N/A	<u>50</u>	<u>4926</u>	mgad/hour
96000832	AN-105	8 DUP	S96T003908 <u>5138g</u> →	0 W		H2001G01	SOLID	<u>4926</u>	<u>5138</u>	N/A g/L
96000832	AN-105	9 DUP	S96T003908 <u>5138g</u> →	0 W	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	mgad/hour

Final page for worklist #

11151

Analyst Signature Rae Ann Green Date 8-5-96Analyst Signature Rae Ann Green Date 8-10-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11152

Analyst: Dr D Instrument: H2001 Book # _____Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN105 C152 SEG16L/18U H20DIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FORMD	DL	UNIT
		1	BLNK-PREP		H2001G01	SOLID	<u>1</u>	<u>1.000</u>	N/A	g/L
96000832	AN-105	2	SAMPLE	8961003909	0 W	H2001G01	SOLID	N/A	<u>5.016</u>	g/L
				<u>.6026 g</u>	<u>→ .1001</u>					
96000832	AN-105	3	SAMPLE	8961003909	0 W	D05E-02	SOLID	N/A	<u>50</u>	mg/d/hour
96000832	AN-105	4	DUP	8961003909	0 W	H2001G01	SOLID	<u>5.016</u>	<u>4.994</u>	N/A
				<u>.4994 g</u>	<u>→ .1001</u>					
96000832	AN-105	5	DUP	8961003909	0 W	D05E-02	SOLID	<u>50</u>	<u>50</u>	N/A
96000832	AN-105	6	SAMPLE	8961003910	0 W	H2001G01	SOLID	N/A	<u>4.912</u>	g/L
				<u>.4912 g</u>	<u>→ .1001</u>					
96000832	AN-105	7	SAMPLE	8961003910	0 W	D05E-02	SOLID	N/A	<u>40</u>	mg/d/hour
96000832	AN-105	8	DUP	8961003910	0 W	H2001G01	SOLID	<u>4.912</u>	<u>5.052</u>	N/A
				<u>.5052 g</u>	<u>→ .1001</u>					
96000832	AN-105	9	DUP	8961003910	0 W	D05E-02	SOLID	<u>40</u>	<u>40</u>	N/A

Final page for worklist #

11152

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11153

Analyst: DGD Instrument: H2001 Book # _____Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN105 C152 SEG18L/20U H2ODIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK - PREP			H2001001	SOL ID	<u>1</u>	<u>.100</u>	N/A	g/L
96000832	AK-105	2 SAMPLE	9961003911	0 W	H2001001	SOL ID	N/A	<u>5.065</u>		g/L
			<u>.5065g</u> → <u>.100L</u>							
96000832	AK-105	3 SAMPLE	9961003911	0 W	DOSE-02	SOL ID	N/A	<u>50</u>		mrads/hour
96000832	AK-105	4 DUP	9961003911	0 W	H2001001	SOL ID	<u>5.065</u>	<u>5.154</u>	N/A	g/L
			<u>.5154g</u> → <u>.100L</u>							
96000832	AK-105	5 DUP	9961003911	0 W	DOSE-02	SOL ID	<u>50</u>	<u>40</u>	N/A	mrads/hour
96000832	AK-105	6 SAMPLE	9961003912	0 W	H2001001	SOL ID	N/A	<u>4.992</u>		g/L
			<u>.4992g</u> → <u>.100L</u>							
96000832	AK-105	7 SAMPLE	9961003912	0 W	DOSE-02	SOL ID	N/A	<u>40</u>		mrads/hour
96000832	AK-105	8 DUP	9961003912	0 W	H2001001	SOL ID	<u>4.992</u>	<u>4.811</u>	N/A	g/L
			<u>.4811g</u> → <u>.100L</u>							
96000832	AK-105	9 DUP	9961003912	0 W	DOSE-02	SOL ID	<u>40</u>	<u>20</u>	N/A	mrads/hour

Final page for worklist #

11153

Analyst Signature DGD Date 8-5-96Analyst Signature [Signature] Date 8-11-96

Bible Coon 8/6/96

.5961003995 → .3911
.3996 → .3912Reviewed by RK Fuller
8/12/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11154

Analyst: DPB Instrument: H2001 Book # _____

Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN105 C152 SEG20L/20U H2ODIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK. PREP			H2001G01	SOL ID	<u>1</u>	<u>0.100</u>	<u>N/A</u>	<u>g/L</u>
96000832	AN-105	2 SAMPLE	596T003913 <u>0 W</u>		H2001G01	SOL ID	<u>N/A</u>	<u>5.021</u>		<u>g/L</u>
96000832	AN-105	3 SAMPLE	596T003913 <u>0 W</u>		DOSE-02	SOL ID	<u>N/A</u>	<u>445</u>		<u>mrad/hour</u>
96000832	AN-105	4 DUP	596T003913 <u>0 W</u>		H2001G01	SOL ID	<u>5.021</u>	<u>4.627</u>	<u>N/A</u>	<u>g/L</u>
96000832	AN-105	5 DUP	596T003913 <u>0 W</u>		DOSE-02	SOL ID	<u>45</u>	<u>30</u>	<u>N/A</u>	<u>mrad/hour</u>
96000832	AN-105	6 SAMPLE	596T003914 <u>0 W</u>		H2001G01	SOL ID	<u>N/A</u>	<u>4.574</u>		<u>g/L</u>
96000832	AN-105	7 SAMPLE	596T003914 <u>0 W</u>		DOSE-02	SOL ID	<u>N/A</u>	<u>30</u>		<u>mrad/hour</u>
96000832	AN-105	8 DUP	596T003914 <u>0 W</u>		H2001G01	SOL ID	<u>4.574</u>	<u>4.913</u>	<u>N/A</u>	<u>g/L</u>
96000832	AN-105	9 DUP	596T003914 <u>0 W</u>		DOSE-02	SOL ID	<u>30</u>	<u>35</u>	<u>N/A</u>	<u>mrad/hour</u>

Final page for worklist #

11154

Daniel P. Bronley 8/5/96
Analyst Signature Date

[Signature] 8/11/96
Analyst Signature Date

596T003913 → 3913
3798 → 3914

Data Entry Comments:

Reviewed by RK Feller
8/12/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12025

Analyst: ADD Instrument: H2001 Book # _____

Method: LA-504-101 Rev/Mod EO

Worklist Comment: AN105, S96T3981-> S96T3990, S96T3982-> S96T3991 skm

GRAB	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP		H2001G01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L
96000945	AR-105	2	SAMPLE S96T003990	0 W	H2001G01	SOLID	N/A	<u>5.092</u>		g/L
			<u>5092</u> →							
96000945	AR-105	3	SAMPLE S96T003990	0 W	DOSE-02	SOLID	N/A	<u>50</u>		cpad/hour
96000945	AR-105	4	DUP S96T003990	0 W	H2001G01	SOLID	<u>5.092</u>	<u>5.062</u>	N/A	g/L
			<u>5062</u> →							
96000945	AR-105	5	DUP S96T003990	0 W	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	cpad/hour
96000945	AR-105	6	SAMPLE S96T003991	0 W	H2001G01	SOLID	N/A	<u>5.091</u>		g/L
			<u>5091</u> →							
96000945	AR-105	7	SAMPLE S96T003991	0 W	DOSE-02	SOLID	N/A	<u>35</u>		cpad/hour
96000945	AR-105	8	DUP S96T003991	0 W	H2001G01	SOLID	<u>5.091</u>	<u>5.063</u>	N/A	g/L
			<u>5063</u> →							
96000945	AR-105	9	DUP S96T003991	0 W	DOSE-02	SOLID	<u>35</u>	<u>30</u>	N/A	cpad/hour

Final page for worklist # 12025

Kim Thomas 8-26-96
Analyst Signature Date

Tim O'Leary 08-22-96
Analyst Signature Date

Kim Thomas 8-26-96

Data Entry Comments:

UPT-Scott Corbally Reviewed by RK Fuller
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12026

Analyst: AD Instrument: H2001 Book # _____

Method: LA-504-101 Rev/Mod E O

Worklist Comment: AN105, H2ODIG01, S96T3983-> S96T3992 skm

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2ODIG01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L
96000945	AR-105	2 SAMPLE	S96T003992	0 W	H2ODIG01	SOLID	N/A	<u>5.318</u>		g/L
96000945	AR-105	3 SAMPLE	S96T003992	0 W	DOSE-02	SOLID	N/A	<u>50</u>		mrad/hour
96000945	AR-105	4 DUP	S96T003992	0 W	H2ODIG01	SOLID	<u>5.318</u>	<u>5.318</u>	N/A	g/L
96000945	AR-105	5 DUP	S96T003992	0 W	DOSE-02	SOLID	<u>450</u>	<u>45</u>	N/A	mrad/hour

Final page for worklist # 12026

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

12113

Analyst: ADP Instrument: H2001 Book #

Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN-105 H2O digest C153 Seg 14 LH, 15 UH lad

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT	
		1	BLNK - PREP		H2001G01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L	
96000945	AN-105	2	SAMPLE • 5365 →	0	W	H2001G01	SOLID	N/A	<u>5.365</u>	g/L	
96000945	AN-105	3	SAMPLE	0	W	DOSE-02	SOLID	N/A	<u>60</u>	mg/d/hour	
96000945	AN-105	4	DUP • 5440 →	0	W	H2001G01	SOLID	<u>5.365</u>	<u>5.440</u>	N/A	g/L
96000945	AN-105	5	DUP	0	W	DOSE-02	SOLID	<u>60</u>	<u>60</u>	N/A	mg/d/hour
96000945	AN-105	6	SAMPLE • 5263 →	0	W	H2001G01	SOLID	N/A	<u>5.263</u>	g/L	
96000945	AN-105	7	SAMPLE	0	W	DOSE-02	SOLID	N/A	<u>60</u>	mg/d/hour	
96000945	AN-105	8	DUP • 5260 →	0	W	H2001G01	SOLID	<u>5.263</u>	<u>5.260</u>	N/A	g/L
96000945	AN-105	9	DUP	0	W	DOSE-02	SOLID	<u>60</u>	<u>60</u>	N/A	mg/d/hour

Final page for worklist #

12113

Anthony Puro 8-26-96
Analyst Signature Date

Richard J. Smith 8/27/96
Analyst Signature Date

Kim Thomas 8-26-96

Data Entry Comments: HPT-Scout Carbally
Reviewed by R.K. Miller
8/28/96

LABCORE Data Entry Template for Worklist#

12114

Analyst: ADD Instrument: H2001 Book #

Method: LA-504-101 Rev/Mod E0

Worklist Comment: AN-105 H2O digest C153 Seg 15 LH, 17 UH lad

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2001001	SOLID	<u>1</u>	<u>100</u>	<u>N/A</u>	<u>g/L</u>
96000945	AN-105	2 SAMPLE <u>5118</u> →	9961004319	0 W	H2001001	SOLID	<u>N/A</u>	<u>5.118</u>		<u>g/L</u>
96000945	AN-105	3 SAMPLE	9961004319	0 W	DOSE-02	SOLID	<u>N/A</u>	<u>60</u>		<u>brad/hour</u>
96000945	AN-105	4 DUP <u>5037</u> →	9961004319	0 W	H2001001	SOLID	<u>5.037</u>	<u>5.001</u>	<u>N/A</u>	<u>g/L</u>
96000945	AN-105	5 DUP	9961004319	0 W	DOSE-02	SOLID	<u>60</u>	<u>60</u>	<u>N/A</u>	<u>brad/hour</u>
96000945	AN-105	6 SAMPLE <u>5210</u> →	9961004447	0 W	H2001001	SOLID	<u>N/A</u>	<u>5.210</u>		<u>g/L</u>
96000945	AN-105	7 SAMPLE	9961004447	0 W	DOSE-02	SOLID	<u>N/A</u>	<u>60</u>		<u>brad/hour</u>
96000945	AN-105	8 DUP <u>5294</u> →	9961004447	0 W	H2001001	SOLID	<u>5.210</u>	<u>5.294</u>	<u>N/A</u>	<u>g/L</u>
96000945	AN-105	9 DUP	9961004447	0 W	DOSE-02	SOLID	<u>60</u>	<u>60</u>	<u>N/A</u>	<u>brad/hour</u>

Final page for worklist #

12114

Altha Purni 8-26-96
Analyst Signature Date

Reinhold 8/27/96
Analyst Signature Date

Rita Thomas 8-26-96

Data Entry Comments:

HPT - SCOTT COMPANY

Reviewed by RkZilla
8/28/96

LABCORE Data Entry Template for Worklist#

12115

Analyst: ADP Instrument: H2001 Book # _____

Method: LA-504-101 Rev/Mod E0

Worklist Comment: AN-105 H2O digests C153 Seg 17 LH, 19 LH had

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLANK-PREP		H2001G01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L
96000945	AN-105	2	SAMPLE 8961004445	0 W	H2001G01	SOLID	N/A	<u>5257</u>		g/L
96000945	AN-105	3	SAMPLE 8961004448	0 W	DOSE-02	SOLID	N/A	<u>65</u>		mgad/hour
96000945	AN-105	4	DUP 8961004448	0 W	H2001G01	SOLID	<u>5257</u>	<u>5257</u>	N/A	g/L
96000945	AN-105	5	DUP 8961004448	0 W	DOSE-02	SOLID	<u>65</u>	<u>65</u>	N/A	mgad/hour
96000945	AN-105	6	SAMPLE 8961004449	0 W	H2001G01	SOLID	N/A	<u>5173</u>		g/L
96000945	AN-105	7	SAMPLE 8961004449	0 W	DOSE-02	SOLID	N/A	<u>65</u>		mgad/hour
96000945	AN-105	8	DUP 8961004449	0 W	H2001G01	SOLID	<u>5173</u>	<u>5173</u>	N/A	g/L
96000945	AN-105	9	DUP 8961004449	0 W	DOSE-02	SOLID	<u>65</u>	<u>65</u>	N/A	mgad/hour

Final page for worklist #

12115

Kim Thomas 8-26-96
Analyst Signature Date

Kim Thomas 8-26-96
Analyst Signature Date

Data Entry Comments:

Reviewed by RR Fulk
HPT-SCOTT COMPANY 8/28/96

LABCORE Data Entry Template for Worklist# 12253

Analyst: _____ Instrument: H2001 _____ Book # _____

Method: LA-504-101 Rev/Mod _____

Worklist Comment: HEY, H2ODIG01, S96R000549 -> S96R000593 skm

GROUP	PROJECT	S TYPE	SAMPLER#	R A	TEST1	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2001G01	SOLID	1	100	N/A	g/L
96000995	HEY	2 SAMPLE	S96R000593 0 W		H2001G01	SOLID	N/A	5.360		g/L
			• 536g → • 100 mL							
96000995	HEY	3 SAMPLE	S96R000593 0		DOSE-02	SOLID	N/A	365		arad/hour
96000995	HEY	4 DUP	S96R000593 0 W		H2001G01	SOLID	5.360	5.061	N/A	g/L
			• 5061g → • 100 mL							
96000995	HEY	5 DUP	S96R000593 0		DOSE-02	SOLID	365	270	N/A	arad/hour

Final page for worklist # 12253

Analyst Signature: *Anthony Peruto* Date: 8-25-96

Analyst Signature: *21162* Date: 8-26-96

Analyst Signature: *Anthony Peruto* Date: 8-24-96

Analyst Signature: *Dave Del* Date: 8-25-96

Reviewed by *PK Fuller*
8/27/96

Data Entry Comments:

HPT - SAM SHEMA

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

27A

LABCORE Data Entry Template for Worklist# 12385

Analyst: DPB Instrument: H2001 Book #

Method: LA-504-101 Rev/Mod E-0

Worklist Comment: AN-105 H2O Digest Seg 14-20 Core Comp lad and Seg 22 Core Comp ~~154086~~

GRUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2001001	SOLID	<u>1</u>	<u>100</u>	N/A	g/L
96000832	AN-105	2 SAMPLE	5967004258	G W	H2001001	SOLID	N/A	<u>4.7030</u>		g/L
96000832	AN-105	3 SAMPLE	5967004258	G W	DOSE-02	SOLID	N/A	<u>40</u>		brad/hour
96000832	AN-105	4 DUP	5967004258	G W	H2001001	SOLID	<u>4.7030</u>	<u>4.8760</u>	N/A	g/L
96000832	AN-105	5 DUP	5967004258	G W	DOSE-02	SOLID	<u>40</u>	<u>40</u>	N/A	brad/hour
96000832	AN-105	6 SAMPLE	5967004263	G W	H2001001	SOLID	N/A	<u>5.3160</u>		g/L
96000832	AN-105	7 SAMPLE	5967004263	G W	DOSE-02	SOLID	N/A	<u>35</u>		brad/hour
96000832	AN-105	8 DUP	5967004263	G W	H2001001	SOLID	<u>5.3160</u>	<u>5.3660</u>	N/A	g/L
96000832	AN-105	9 DUP	5967004263	G W	DOSE-02	SOLID	<u>35</u>	<u>35</u>	N/A	brad/hour

Final page for worklist # 12385

Daniel P. Bishop 9/5/96
Analyst Signature Date

D. Murphy 9-6-96
Analyst Signature Date

5967004254 → 4258 14-20 Comp
5967004259 → 4263 22 Comp

Data Entry Comments:

Submitted by DPB HPT Bill Bishop

Reviewed by RPK
9/9/96

LABCORE Data Entry Template for Worklist# 13659

Analyst: SHE Instrument: H2001 Book # NA

Method: LA-504-101 Rev/Mod E-D

Worklist Comment: AN-105 H2O DIGEST C153S17, 19 LH lad

GROUP	PROJECT	S TYPE	SAMPLE	R X	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREF		4200-1001	SOLID	1	100	N/A	g/L
96000945	AN-105	2	SAMPLE 5424g → 100.8	0 W	4200-1001	SOLID	N/A	5.424		g/L
96000945	AN-105	3	SAMPLE 5424g → 100.8	0 W	0050-02	SOLID	N/A	75		mg/dl/hour
96000945	AN-105	4	DUP 5424g → 100.8	0 W	4200-1001	SOLID	5.424	5.492	N/A	g/L
96000945	AN-105	5	DUP 5424g → 100.8	0 W	0050-02	SOLID	75	75	N/A	mg/dl/hour
96000945	AN-105	6	SAMPLE 5406g → 100.8	0 W	4200-1001	SOLID	N/A	5.406		g/L
96000945	AN-105	7	SAMPLE 5406g → 100.8	0 W	0050-02	SOLID	N/A	75		mg/dl/hour
96000945	AN-105	8	DUP 5406g → 100.8	0 W	4200-1001	SOLID	5.406	5.392	N/A	g/L
96000945	AN-105	9	DUP 5406g → 100.8	0 W	0050-02	SOLID	75	75	N/A	mg/dl/hour

Final page for worklist # 13659

Analyst Signature: Lusie M. Dalton Date: 10-28-96

Analyst Signature: Frank Mark Date: 10-28-96

S96T004436 → S96T004 ^{91F} 10-28-96
S96T005325

S96T004437 → S96T005326

Revised: 10/28/96

Data Entry Comments:

Weighed by SHE, HPT Bill Bishop
all samples - 75 mg/dl/hr / 50 mg/hr

LABCORE Data Entry Template for Worklist#

11148

Analyst: Ded R. Gert Instrument: ACD01 Book # _____Method: LA-505-159 Rev/Mod D-0

Worklist Comment: AN105 C152 SEG1U/14U/14L/16U/16L/18U ACIDIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	1	.050	N/A	g/L
		2 STD-PREP			ACIDIG01	SOLID	20	20	N/A	g/L
96000832	AN-105	3 SAMPLE	S961003906 0 A .2544g →		ACIDIG01 .050	SOLID	N/A	5.088		g/L
96000832	AN-105	4 SAMPLE	S961003906 0 A		DOSE-02	SOLID	N/A	25		nrads/hour
96000832	AN-105	5 DUP	S961003906 0 A .2463g →		ACIDIG01 .050	SOLID	5.138	4.964	N/A	g/L
96000832	AN-105	6 DUP	S961003906 0 A		DOSE-02	SOLID	25	22	N/A	nrads/hour
96000832	AN-105	7 SPK	S961003906 0 A .2463g →		ACIDIG01 .050	SOLID	20	20	N/A	g/L
96000832	AN-105	8 SPK	S961003906 0 A		DOSE-02	SOLID	N/A	20	N/A	nrads/hour
96000832	AN-105	9 SAMPLE	S961003896 0 A .2543g →		ACIDIG01 .050	SOLID	N/A	5.086		g/L
96000832	AN-105	10 SAMPLE	S961003896 0 A		DOSE-02	SOLID	N/A	35		nrads/hour
96000832	AN-105	11 DUP	S961003896 0 A .2577g →		ACIDIG01 .050	SOLID	5.086	5.154	N/A	g/L
96000832	AN-105	12 DUP	S961003896 0 A		DOSE-02	SOLID	35	35	N/A	nrads/hour
96000832	AN-105	13 SAMPLE	S961003828 0 A .2478g →		ACIDIG01 .050	SOLID	N/A	4.950		g/L
96000832	AN-105	14 SAMPLE	S961003828 0 A		DOSE-02	SOLID	N/A	25		nrads/hour
96000832	AN-105	15 DUP	S961003828 0 A .2544g →		ACIDIG01 .050	SOLID	4.950	5.088	N/A	g/L
96000832	AN-105	16 DUP	S961003828 0 A		DOSE-02	SOLID	25	25	N/A	nrads/hour
96000832	AN-105	17 SAMPLE	S961003899 0 A .2523g →		ACIDIG01 .050	SOLID	N/A	5.086		g/L
96000832	AN-105	18 SAMPLE	S961003899 0 A		DOSE-02	SOLID	N/A	30		nrads/hour

Data Entry Comments:

RK fulla
8/12/96Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

2577

LABCORE Data Entry Template for Worklist#

11148

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
96000832	AN-105	19 DUP	596T003899 .2511 g →	G A	ACID1601 0501	SOLID	5046	5042	N/A	g/L
96000832	AN-105	20 DUP	596T003900 0501	G A	DOSE-02	SOLID	30	40	N/A	mrads/hour
96000832	AN-105	21 SAMPLE	596T003900 .2596 g →	G A	ACID1601 0501	SOLID	N/A	5092		g/L
96000832	AN-105	22 SAMPLE	596T003900 0501	G A	DOSE-02	SOLID	N/A	35		mrads/hour
96000832	AN-105	23 DUP	596T003900 .2546 g →	G A	ACID1601 0501	SOLID	5092	5092	N/A	g/L
96000832	AN-105	24 DUP	596T003900 0501	G A	DOSE-02	SOLID	35	40	N/A	mrads/hour
96000832	AN-105	25 SPK	596T003900 .2598 g →	G A	ACID1601 0501	SOLID	20	20	N/A	g/L
96000832	AN-105	26 SPK	596T003900 0501	G A	DOSE-02	SOLID	N/A	30	N/A	mrads/hour
96000832	AN-105	27 SAMPLE	596T003901 .2572 g →	G A	ACID1601 0501	SOLID	N/A	5144		g/L
96000832	AN-105	28 SAMPLE	596T003901 0501	G A	DOSE-02	SOLID	N/A	30		mrads/hour
96000832	AN-105	29 DUP	596T003901 .2524 g →	G A	ACID1601 0501	SOLID	5144	5018	N/A	g/L
96000832	AN-105	30 DUP	596T003901 0501	G A	DOSE-02	SOLID	30	30	N/A	mrads/hour

Final page for worklist #

11148

Analyst Signature

Date

8-5-96

8-11-96

Analyst Signature

Date

8/12/96

296T003731 → 3906
3876 → 3598
3817 → 3828
3792 → 3899
3793 → 3410
3794 → 3901

Data Entry Comments:

Reviewed by RK-Tulla
8/10/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11149

Analyst: BOD KSH Instrument: ACD01 Book # _____

Method: LA-505-159 Rev/Mod DO

Worklist Comment: AN105 C152 SEG18L/20U/20L/22U ACIDIG01 SKB

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP		ACIDIG01	SOLID	<u>1</u>	<u>.080</u>	N/A	g/L
		2	STD-PREP		ACIDIG01	SOLID	<u>40</u>	<u>20</u>	N/A	g/L
96090832	AN-105	3	SAMPLE	S961003902	C A	ACIDIG01	SOLID	N/A	<u>5004</u>	g/L
				<u>.2502g</u> →	<u>.0501</u>					
96090832	AN-105	4	SAMPLE	S961003902	C A	DOSE-U2	SOLID	N/A	<u>30</u>	mrad/hour
96090832	AN-105	5	DUP	S961003902	C A	ACIDIG01	SOLID	<u>5004</u>	<u>4946</u>	N/A
				<u>.2473g</u> →	<u>.0501</u>					
96090832	AN-105	6	DUP	S961003902	C A	DOSE-U2	SOLID	<u>30</u>	<u>30</u>	N/A
96090832	AN-105	7	SPK	S961003902	C A	ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A
				<u>.2464g</u> →	<u>.0501</u>					
96090832	AN-105	8	SPK	S961003902	C A	DOSE-U2	SOLID	N/A	<u>05</u>	mrad/hour
96090832	AN-105	9	SAMPLE	S961003903	C A	ACIDIG01	SOLID	N/A	<u>5222</u>	g/L
				<u>.2611g</u> →	<u>.0501</u>					
96090832	AN-105	10	SAMPLE	S961003903	C A	DOSE-U2	SOLID	N/A	<u>30</u>	mrad/hour
96090832	AN-105	11	DUP	S961003903	C A	ACIDIG01	SOLID	<u>5222</u>	<u>5316</u>	N/A
				<u>.2658g</u> →	<u>.0501</u>					
96090832	AN-105	12	DUP	S961003903	C A	DOSE-U2	SOLID	<u>30</u>	<u>30</u>	N/A
96090832	AN-105	13	SAMPLE	S961003904	C A	ACIDIG01	SOLID	N/A	<u>5330</u>	g/L
				<u>.2665g</u> →	<u>.0501</u>					
96090832	AN-105	14	SAMPLE	S961003904	C A	DOSE-U2	SOLID	N/A	<u>30</u>	mrad/hour
96090832	AN-105	15	DUP	S961003904	C A	ACIDIG01	SOLID	<u>5330</u>	<u>5394</u>	N/A
				<u>.2644g</u> →	<u>.0501</u>					
96090832	AN-105	16	DUP	S961003904	C A	DOSE-U2	SOLID	<u>30</u>	<u>40</u>	N/A
96090832	AN-105	17	SPK	S961003904	C A	ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A
				<u>.2743g</u> →	<u>.0502</u>					
96090832	AN-105	18	SPK	S961003904	C A	DOSE-U2	SOLID	N/A	<u>30</u>	mrad/hour

Data Entry Comments:

*Received by RK Fuller
8/12/96*

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist#

11149

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
96000832	AN-105	19 SAMPLE	996T003905	0 A	ACID1001	SOLID	N/A	5.050		g/l
			2525 g		1000 l					
96000832	AN-105	20 SAMPLE	996T003905	0 A	DOSE-02	SOLID	N/A	40		mgad/hour
96000832	AN-105	21 DUP	996T003905	0 A	ACID1001	SOLID	5.050	4.878	N/A	g/l
			2414 g		1000 l					
96000832	AN-105	22 DUP	996T003905	0 A	DOSE-02	SOLID	40	40	N/A	mgad/hour

Final page for worklist #

11149

David D. [Signature] 8-5-96
Analyst Signature Date

T. [Signature] 08-12-96
Analyst Signature Date

296T003905 → 3902
3796 → 3903
3797 → 3904
3798 → 3905

Data Entry Comments:

Reviewed by P.K. [Signature]
8/1/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

worklist Version 2.1 05/15/95
08/26/96 12:42

LABCORE Data Entry Template for Worklist# 12110

Analyst: Jan Instrument: ACD01 Book # _____Method: LA-S05-159 Rev/Mod B2.00

Worklist Comment: AN-105 Acid digest C153 Seg 15 UH, LH, 14 LH, 17 UH g/b

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	UL	UNIT
		1	STD - PREP				ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
		2	BLANK - PREP				ACID1601	SOLID	<u>1</u>	<u>0.70</u>	N/A	g/L
96000945	AN-105	3	SAMPLE	S96T004315	0	A	ACID1601	SOLID	N/A	<u>5.213</u>		g/L
				<u>2.050</u>								
96000945	AN-105	4	SAMPLE	S96T004315	0	A	DOSE-02	SOLID	N/A	<u>40</u>		mg/dl/hour
96000945	AN-105	5	DUP	S96T004315	0	A	ACID1601	SOLID	<u>5.246</u>	<u>5.225</u>	N/A	g/L
				<u>2.050</u>								
96000945	AN-105	6	DUP	S96T004315	0	A	DOSE-02	SOLID	<u>40</u>	<u>40</u>	N/A	mg/dl/hour
96000945	AN-105	7	SAMPLE	S96T004318	0	A	ACID1601	SOLID	N/A	<u>5.213</u>		g/L
				<u>2.050</u>								
96000945	AN-105	8	SAMPLE	S96T004318	0	A	DOSE-02	SOLID	N/A	<u>50</u>		mg/dl/hour
96000945	AN-105	9	DUP	S96T004318	0	A	ACID1601	SOLID	<u>5.213</u>	<u>5.246</u>	N/A	g/L
				<u>2.050</u>								
96000945	AN-105	10	DUP	S96T004318	0	A	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	mg/dl/hour
96000945	AN-105	11	SPK	S96T004318	0	A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
				<u>2.050</u>								
96000945	AN-105	12	SPK	S96T004318	0	A	DOSE-02	SOLID	<u>NA</u>	<u>50</u>	N/A	mg/dl/hour
96000945	AN-105	13	SAMPLE	S96T004338	0	A	ACID1601	SOLID	N/A	<u>5.354</u>		g/L
				<u>2.050</u>								
96000945	AN-105	14	SAMPLE	S96T004338	0	A	DOSE-02	SOLID	N/A	<u>40</u>		mg/dl/hour
96000945	AN-105	15	DUP	S96T004338	0	A	ACID1601	SOLID	<u>5.354</u>	<u>5.246</u>	N/A	g/L
				<u>2.050</u>								
96000945	AN-105	16	DUP	S96T004338	0	A	DOSE-02	SOLID	<u>40</u>	<u>40</u>	N/A	mg/dl/hour
96000945	AN-105	17	SPK	S96T004338	0	A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
				<u>2.050</u>								
96000945	AN-105	18	SPK	S96T004338	0	A	DOSE-02	SOLID	<u>NA</u>	<u>40</u>	N/A	mg/dl/hour

Data Entry Comments:

HPT-Bill Bishop

Reviewed by RKJ
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12110

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FORM	DL	UNIT
96000945	AN-105	19 SAMPLE	596T004444	0 A	ACID1001	SOLID	N/A	5.174		g/L
		12587								
96000945	AN-105	20 SAMPLE	596T004444	0 A	DOSE-02	SOLID	N/A	40		mgad/hour
96000945	AN-105	21 DUP	596T004444	0 A	ACID1001	SOLID	5.174	5.204	N/A	g/L
		2432								
96000945	AN-105	22 DUP	596T004444	0 A	DOSE-02	SOLID	40	40	N/A	mgad/hour

Final page for worklist # 12110

Analyst Signature: [Signature] Date: 8-26-96

Analyst Signature: [Signature] Date: 8-26-96

T.R. Cheron 08-27-96

Data Entry Comments:

Reviewed by RK Fulla
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Worklist Version 2.1 05/15/95
08/26/96 13:05

Page: 1

LABCORE Data Entry Template for Worklist# 12111

Analyst: AKM Instrument: ACD01 Book # _____

Method: LA-505-159 Rev/Mod 88 DO

Worklist Comment: AN-105 Acid Digest C153 Seg 21 LH, 17 Lh, 19 LH gjh

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD-PREP				ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
		2	BLNK-PREP				ACID1601	SOLID	<u>1</u>	<u>0.50</u>	N/A	g/L
96000945	AN-105	3	SAMPLE	596T003988	0	A	ACID1601	SOLID	N/A	<u>5.124</u>		g/L
96000945	AN-105	4	SAMPLE	596T003988	0	A	DOSE-02	SOLID	N/A	<u>50</u>		mg/d/hour
96000945	AN-105	5	DUP	596T003988	0	A	ACID1601	SOLID	<u>5.124</u>	<u>5.122</u>	N/A	g/L
96000945	AN-105	6	DUP	596T003988	0	A	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	mg/d/hour
96000945	AN-105	7	SPK	596T004445	0	A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
96000945	AN-105	8	SPK	596T004445	0	A	DOSE-02	SOLID	N/A	<u>50</u>		mg/d/hour
96000945	AN-105	9	SAMPLE	596T004445	0	A	ACID1601	SOLID	N/A	<u>5.230</u>		g/L
96000945	AN-105	10	SAMPLE	596T004445	0	A	DOSE-02	SOLID	N/A	<u>50</u>		mg/d/hour
96000945	AN-105	11	DUP	596T004445	0	A	ACID1601	SOLID	<u>5.230</u>	<u>5.232</u>	N/A	g/L
96000945	AN-105	12	DUP	596T004445	0	A	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	mg/d/hour
96000945	AN-105	13	SAMPLE	596T004446	0	A	ACID1601	SOLID	N/A	<u>5.230</u>		g/L
96000945	AN-105	14	SAMPLE	596T004446	0	A	DOSE-02	SOLID	N/A	<u>50</u>		mg/d/hour
96000945	AN-105	15	DUP	596T004446	0	A	ACID1601	SOLID	<u>5.230</u>	<u>5.232</u>	N/A	g/L
96000945	AN-105	16	DUP	596T004446	0	A	DOSE-02	SOLID	<u>50</u>	<u>50</u>	N/A	mg/d/hour

Data Entry Comments:

HPT-Bill Bishop

Reviewed by CR Fuller
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12111

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
-------	---------	--------	---------	-----	------	--------	--------	-------	----	------

Final page for worklist # 12111

[Signature] 8-26-96
Analyst Signature Date

[Signature] 8-26-96
Analyst Signature Date

TR. Olsen 8-27-96

Data Entry Comments:

RK Fuller 8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12112

Analyst: RAM Instrument: ACD01 Book # Method: LA-505-159 Rev/Mod 82-00

Worklist Comment: AN-105 Acid digest C153 Seg 22 LH UH g/h

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD-PREP			ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
		2 BLNK-PREP			ACID1601	SOLID	<u>1</u>	<u>150</u>	N/A	g/L
9600945	AN-105	3 SAMPLE	996T003987	0 A	ACID1601	SOLID	N/A	<u>5.132</u>		g/L
			<u>2560</u> → <u>7.050</u> L							
9600945	AN-105	4 SAMPLE	996T003987	0 A	DOSE-02	SOLID	N/A	<u>80</u>		mg/d/hour
9600945	AN-105	5 DUP	996T003987	0 A	ACID1601	SOLID	<u>5.132</u>	<u>5.141</u>	N/A	g/L
			<u>2557</u> → <u>7.050</u> L							
9600945	AN-105	6 DUP	996T003987	0 A	DOSE-02	SOLID	<u>80</u>	<u>80</u>	N/A	mg/d/hour
9600945	AN-105	7 SPK	996T003987	0 A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
			<u>2547</u> → <u>7.050</u> L							
9600945	AN-105	8 SPK	996T003987	0 A	DOSE-02	SOLID	N/A	<u>80</u>	N/A	mg/d/hour
9600945	AN-105	9 SAMPLE	996T003989	0 A	ACID1601	SOLID	N/A	<u>5.374</u>		g/L
			<u>2687</u> → <u>7.050</u> L							
9600945	AN-105	10 SAMPLE	996T003989	0 A	DOSE-02	SOLID	N/A	<u>40</u>		mg/d/hour
9600945	AN-105	11 DUP	996T003989	0 A	ACID1601	SOLID	<u>5.374</u>	<u>5.442</u>	N/A	g/L
			<u>2706</u> → <u>7.050</u> L							
9600945	AN-105	12 DUP	996T003989	0 A	DOSE-02	SOLID	<u>40</u>	<u>40</u>	N/A	mg/d/hour

Final page for worklist # 12112

Analyst Signature RAM Date 8-30-96Analyst Signature [Signature] Date 8-26-96Data Entry Comment: HPT-Bird BishopRK Folley
8/28/96Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12252

Analyst: ADP Instrument: ACD01 Book # _____

Method: LA-505-159 Rev/Mod BA DO

Worklist Comment: HEY, ACIDIG01, S96R000549-> S96R000592 skm

GROUP	PRIORITY	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT	
		1	BLNK-PREP		ACIDIG01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L	
		2	STD-PREP		ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A	g/L	
96000995	HEY	3	SAMPLE	S96R000592	0 A	ACIDIG01	SOLID	N/A	<u>5108</u>	g/L	
				2564 →	7.0508						
96000995	HEY	4	SAMPLE	S96R000592	0	D086-02	SOLID	N/A	<u>100</u>	mgad/hour	
96000995	HEY	5	DUP	S96R000592	0 A	ACIDIG01	SOLID	<u>5108</u>	<u>5108</u>	N/A	g/L
				2580 →	0.0508						
96000995	HEY	6	DUP	S96R000592	0	D086-02	SOLID	<u>100</u>	<u>100</u>	N/A	mgad/hour
96000995	HEY	7	SPK	S96R000592	0 A	ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
				2572 →	0.0508						
96000995	HEY	8	SPK	S96R000592	0	D086-02	SOLID	<u>NA</u>	<u>100</u>	N/A	mgad/hour

Final page for worklist # 12252

Analyst Signature: ADP Date: 8-24-96
GM Murphy 8-26-96

Analyst Signature: [Signature] Date: 8-26-96
[Signature]
T.R. Olson 8-27-96

Data Entry Comments: APT-Bill Bishop Reviewed by PKHalla
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units DL = Detection Limit, S = Worklist Shot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12264

Analyst: ADD Instrument: ACD01 Book # _____

Method: LA-505-158 Rev/Mod B-D D O

Worklist Comment: AN105, ACIDIG02, S96T004144 -> S96T004144 skm

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLANK-PREP				ACIDIG02	LIQUID	<u>1</u>	<u>0.00</u>	N/A	DF
		2	STD-PREP				ACIDIG02	LIQUID	<u>20</u>	<u>20</u>	N/A	DF
96000832	AN-105	3	SAMPLE	S96T004144	0	B	ACIDIG02	LIQUID	N/A	<u>5.308</u>		DF
				<u>.2651</u>								
96000832	AN-105	4	SAMPLE	S96T004144	0	B	DOSE-02	LIQUID	N/A	<u>50</u>		grad/hour
96000832	AN-105	5	DUP	S96T004144	0	B	ACIDIG02	LIQUID	<u>5.308</u>	<u>5.598</u>	N/A	DF
				<u>12779</u>								
96000832	AN-105	6	DUP	S96T004144	0	B	DOSE-02	LIQUID	<u>50</u>	<u>50</u>	N/A	grad/hour
96000832	AN-105	7	SPK	S96T004144	0	B	ACIDIG02	LIQUID	<u>20</u>	<u>20</u>	N/A	DF
				<u>2764</u>								
96000832	AN-105	8	SPK	S96T004144	0	B	DOSE-02	LIQUID	N/A	<u>50</u>	N/A	grad/hour

Final page for worklist # 12264

Analyst Signature [Signature] Date 8-24-96
[Signature] 8-26-96

Analyst Signature [Signature] Date 8-26-96
[Signature] 8/27/96

Data Entry Comments

ART-Bill Bishop

Reviewed by PK Fuller
8/28/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12384

Analyst: _____ Instrument: ACD01 ELU01 Book # _____Method: I.A-505-159 Rev/Mod DOWorklist Comment: AN-105 Acid Digest Seg 14-20 Core Comp lad and Seg 22 Core Comp OK

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP		ACID1601	SOLID	<u>1</u>	<u>.050</u>	N/A	g/L
		2	STD-PREP		ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
96000832	AN-105	3	SAMPLE	S96T004257	0 A	ACID1601	SOLID	N/A	<u>5.2240</u>	g/L
				<u>.2412g</u> → <u>.0508</u>						
96000832	AN-105	4	SAMPLE	S96T004257	0 A	DOSE-02	SOLID	N/A	<u>25</u>	mg/dl/hour
96000832	AN-105	5	DUP	S96T004257	0 A	ACID1601	SOLID	<u>5.2240</u>	<u>4.9620</u>	N/A
				<u>.2481g</u> → <u>.0508</u>						
96000832	AN-105	6	DUP	S96T004257	0 A	DOSE-02	SOLID	<u>25</u>	<u>25</u>	N/A
										mg/dl/hour
* 96000832	AN-105	7	SPK	S96T004257	0 A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A
				<u>.2584g</u> → <u>.0508</u>						g/L
96000832	AN-105	8	SPK	S96T004257	0 A	DOSE-02	SOLID	N/A	<u>40</u>	mg/dl/hour
96000832	AN-105	9	SAMPLE	S96T004262	0 A	ACID1601	SOLID	N/A	<u>5.1360</u>	g/L
				<u>.2588g</u> → <u>.0508</u>						
96000832	AN-105	10	SAMPLE	S96T004262	0 A	DOSE-02	SOLID	N/A	<u>40</u>	mg/dl/hour
96000832	AN-105	11	DUP	S96T004262	0 A	ACID1601	SOLID	<u>5.1360</u>	<u>5.4000</u>	N/A
				<u>.2703g</u> → <u>.0508</u>						
96000832	AN-105	12	DUP	S96T004262	0 A	DOSE-02	SOLID	<u>40</u>	<u>30</u>	N/A
										mg/dl/hour
* 96000832	AN-105	13	SPK	S96T004262	0 A	ACID1601	SOLID	<u>20</u>	<u>20</u>	N/A
				<u>.2589g</u> → <u>.0508</u>						g/L
96000832	AN-105	14	SPK	S96T004262	0 A	DOSE-02	SOLID	N/A	<u>40</u>	mg/dl/hour

Final page for worklist # 12384

Analyst Signature Chammying Date 9-6-96Analyst Signature [Signature] Date 9-8-96

Data Entry Comments:

S96T004254 → 4257

4259 → 4262

HPT Dil Bishop

Washed out, kept fresh B. Simon 9/6/96

<1% Solids

Submitted 9/6/96 Edm

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

288

Reviewed by R.K. Field
9/9/96

HNF-SD-WM-OP-199, REV. 1

WHC-SD-WM-OP-199, REV. 1

AN-106 BULK DENSITY WORKSHEETS

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

HNF-SO-WM-DP-199, REV. 1

Bulk Density Worksheet

Requestor: F. SpeedTank: A4105Core: 152Seg: 144Auger: 144Sample ID: 5967003744Start Time: 16:00End Time: 16:00Homogenization Time (Min): 16:00Jar#: 10343Jar/Vial Size: 60 mLInitial Weight: 25.6 gFinal Weight: 17.9 gNet Weight: 7.7 gConc#: 43Final Vol: 11.8 mLInitial Weight: 7.7 gFinal Weight: 25.6 gNet Weight: 17.9 gSample ID: 5967003744Appearance/Narrative: 17.9/11.8Bulk Density: 1.52Date: 7-27-96Tank: 144Core: 144Seg: 144Auger: 144Sample ID: 5967003744Start Time: 16:00End Time: 16:00Homogenization Time (Min): 16:00Jar#: 10343Jar/Vial Size: 60 mLInitial Weight: 25.6 gFinal Weight: 17.9 gNet Weight: 7.7 gConc#: 43Final Vol: 11.8 mLInitial Weight: 7.7 gFinal Weight: 25.6 gNet Weight: 17.9 gSample ID: 5967003744Appearance/Narrative: 17.9/11.8Bulk Density: 1.52

Requestor: _____

Bulk Density Worksheet

Date: 7-26-96Tank: AN-105Core: 152Seg: 14 LH

Auger: _____

Sample ID: _____

Start Time: _____
End Time: _____
Homogenization
Time (Min): _____Jar#: 10394

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

282

Conc#: 94Final Vol: 102.0 mLInitial Weight: 7.69 gFinal Weight: 23.87 gNet Weight: 150.58 g

Sample ID: _____

Appearance/Narrative:

BKDN01 = 15.58 / 10.20 = 1.53 g/mLTank: AN-105Core: 152Seg: 14 LH

Auger: _____

Sample ID: _____

Start Time: _____
End Time: _____
Homogenization
Time (Min): _____Jar#: 10392

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Conc#: 92Final Vol: 10.50 mLInitial Weight: 7.70 gFinal Weight: 23.49 gNet Weight: 157.74 g

Sample ID: _____

Appearance/Narrative:

BKDN01 = 15.74 / 10.50 = 1.50 g/mL

Bulk Density Worksheet

Requestor: _____

Date: 7-26-96

Tank: AN-105

Core: 152

Seg: 164H

Auger: _____

Sample ID: _____

Tank: AN-105

Core: 152

Seg: 164H

Auger: _____

Sample ID: _____

Jar#: 10350

Jar/Vial Size: _____ mL

Initial Weight: 7.53 g

Final Weight: 7.53 g

Net Weight: 0 g

Cone#: 50

Final Vol: 110.0 mL

Initial Weight: 7.53 g

Final Weight: 7.53 g

Net Weight: 0 g

Sample ID: _____

Appearance/Narrative: _____

SLK DENWIS 17.53/110.0 = 1.589 g/mL

Jar#: 10393

Jar/Vial Size: _____ mL

Initial Weight: _____

Final Weight: _____

Net Weight: _____

Cone#: 93

Final Vol: 10.50 mL

Initial Weight: 7.51 g

Final Weight: 24.14 g

Net Weight: 16.63 g

Sample ID: _____

Appearance/Narrative: 10.50

SLK DENWIS 16.63/10.50 = 1.589 g/mL

7.2690

HNF-SD-WM-OP-199, REV. 1

No. 3426

10/18

LC 96
7.53 g
7.53 g

7.2690

1.589 g/mL

Bulk Density Worksheet

Requestor: F. S. M. N.

Date: 7/27/96

Tank: AN-105

Core: 152

Seg: 13 UH

Auger: _____

Sample ID: _____

Start Time: _____
End Time: _____
Homogenization
Time (Min.): _____

Jar#: 10409

Jar/Vial Size: 125 mL

Initial Weight: 4 g

Final Weight: 24.3 g

Net Weight: 16.6 g

284

Cone#: 09

Final Vol: 10.0 mL

Initial Weight: 7.7 g

Final Weight: 24.3 g

Net Weight: 16.6 g

Sample ID: 5967003707

Appearance/Narrative: _____

Bulk Density = 1.66 g/cc

Jar#: 10524

Jar/Vial Size: 250 mL

Initial Weight: 9 g

Final Weight: 33.8 g

Net Weight: 24.8 g

Cone#: 24

Final Vol: 11.0 mL

Initial Weight: 7.7 g

Final Weight: 25.8 g

Net Weight: 18.1 g

Sample ID: 5967003708

Appearance/Narrative: _____

Bulk Density = 1.64 g/cc

HNFSO-WM-OP-199, REV. 1

No. 3425 P. 13/18

Bulk Density Worksheet

Requestor: F. Spaul

Date: 7-27-96

Tank: AN 105
Core: 152
Seg: 20 LH
Auger: _____
Sample ID: _____

Tank: AN 105
Core: 152
Seg: 20 LH
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization
Time (Min.): _____

Jar#: 10357
Jar/Vial Size: 250 mL
Initial Weight: 23.5 g
Final Weight: 7.7 g
Net Weight: _____ g

Cone#: 57
Final Vol: 10.0 mL
Initial Weight: 7.7 g
Final Weight: 23.5 g
Net Weight: 15.8 g

Sample ID: SP000370

Appearance/Narrative:

Net Wt = 15.8 / 10.0 = 1.58 g/cc

Start Time: _____
End Time: _____
Homogenization
Time (Min.): _____

Jar#: 10358
Jar/Vial Size: 250 mL
Initial Weight: 24.9 g
Final Weight: 7.7 g
Net Weight: _____ g

Cone#: 58
Final Vol: 10.8 mL
Initial Weight: 7.7 g
Final Weight: 24.9 g
Net Weight: 17.2 g

Sample ID: _____

Appearance/Narrative:

Net Wt = 17.2 / 10.8 = 1.60 g/cc

HNFS-D-WM-DP-100, REV. 1

NA 3436 P. 16/19

Bulk Density Worksheet

Jul 27, 1998 3:01PM

WHO 2225 LAB ROOM 2F BACKSIDE

No. 3426 18/18

HNF-SD-WM-DP-199, REV. 1

Requestor: EG/AN

Tank: AN-105

Core: 152

Seg: 220H

Auger: _____

Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10398
Jar/Vial Size: 125 mL
Initial Weight: 22.1 g
Final Weight: 22.1 g
Net Weight: _____ g

Cone#: 83
Final Vol: 10.8 mL
Initial Weight: 7.7 g
Final Weight: 35.4 g
Net Weight: 17.7 g

Sample ID: 586T003712

Appearance/Narrative: AN-105
AN-105 = 1.64 g/cc

296

Date: 7/27/98

Tank: _____
Core: _____
Seg: _____
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: _____
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Cone#: _____
Final Vol: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

Bulk Density Worksheet

Requestor: GreenDate: 8/17/96

Tank: 46/45
 Core: 152
 Seg: Comp
 Auger: _____
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: 10972
 Jar/Vial Size: _____ mL
 Initial Weight: 16.5 g
 Final Weight: 17.5 g
 Net Weight: _____ g

Cone#: 72
 Final Vol: 10.5 mL
 Initial Weight: 25.0 g
 Final Weight: 17.5 g
 Net Weight: _____ g
 Sample ID: 59670 4208

Appearance/Narrative:

bulk = 1.67 g/cc

Tank: _____
 Core: _____
 Seg: _____
 Auger: _____
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: _____
 Jar/Vial Size: _____ mL
 Initial Weight: _____ g
 Final Weight: 17.5 g
 Net Weight: 19.6 g

Cone#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative:

HNF-SD-WM-DP-199, REV. 1

HNF-SD-WM-OP-199, REV. 1

Bulk Density Worksheet

Requestor: STENDate: 8/14/96Tank: AM-105Core: 151Seg: 14 L H

Auger: _____

Sample ID: _____

Start Time: _____

End Time: _____

Homogenization

Time (Min.): _____

Jar#: 10603

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: 03Final Vol: 10.6 mLInitial Weight: 7.5 gFinal Weight: 24.0 gNet Weight: 16.5 g Sample ID: 5267204334

Appearance/Narrative:

Bulk Density = 1.5 / 10.61.56

Tank: _____

Core: _____

Seg: _____

Auger: _____

Sample ID: _____

Start Time: _____

End Time: _____

Homogenization

Time (Min.): _____

Jar#: _____

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: 12-14

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative:

Bulk Density Worksheet

Date: 8/9/96

HNF-SD-WM-OP-100, REV. 1

Requestor: F. STEEN

Tank: AN-105
Core: 153
Seg: 15UH
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10624
Jar/Vial Size: 250 mL
Initial Weight: 424.5 g
Final Weight: _____ g
Net Weight: _____ g

260

Cone#: 24
Final Vol: 110 mL
Initial Weight: 7.58 g
Final Weight: 24.57 g
Net Weight: 16.96 g

Sample ID: _____

Appearance/Narrative: _____

16.96/11 = 1.52 g/mL

Tank: AN-105
Core: 153
Seg: 15UH
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10603
Jar/Vial Size: 250 mL
Initial Weight: 424.5 g
Final Weight: _____ g
Net Weight: _____ g

Cone#: 03
Final Vol: 120 mL
Initial Weight: 7.47 g
Final Weight: 23.11 g
Net Weight: 15.64 g

Sample ID: _____

Appearance/Narrative: _____

15.64/100 = 1.56 g/mL

HNF-SD-WM-DP-199, REV. 1

Bulk Density Worksheet

Requestor: Steel

Date: 8/19/96

Tank: A-105
Core: 153
Seg: 17th
Auger: ---
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10629
Jar/Vial Size: 1 mL
Initial Weight: 27.5 g
Final Weight: 24.3 g
Net Weight: 3.2 g

Cone#: 29
Final Vol: 11.0 mL
Initial Weight: 27.5 g
Final Weight: 24.3 g
Net Weight: 3.2 g
Sample ID: 576704427

Appearance/Narrative: 153 gpc
153/11.0

Jar#: 10799
Jar/Vial Size: 1 mL
Initial Weight: 27.5 g
Final Weight: 21.4 g
Net Weight: 6.1 g

Cone#: 99
Final Vol: 9.8 mL
Initial Weight: 27.5 g
Final Weight: 21.4 g
Net Weight: 6.1 g
Sample ID: _____

Appearance/Narrative: 152 gpc
152/9.8

Handwritten notes: 4-1-6

HNFS-SD-WM-DP-199, REV. 1

Bulk Density Worksheet

Requestor: 57161

Date: 8/19/96

Tank: A3107
Core: 153
Seg: 19.14
Auger: -
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10625
Jar/Vial Size: 100 mL
Initial Weight: 10.16 g
Final Weight: 8.14 g
Net Weight: 2.02 g

Cone#: 25
Final Vol: 11.5 mL
Initial Weight: 7.5 g
Final Weight: 25.4 g
Net Weight: 17.9 g
Sample ID: 57161004429

Appearance/Narrative: _____

17.9 / 11.5 = 1.56 g/cc

Tank: _____
Core: _____
Seg: _____
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: _____
Jar/Vial Size: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Cone#: _____
Final Vol: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g
Sample ID: _____

Appearance/Narrative: _____

HNF-SD-WM-DP-199, REV. 1

Bulk Density Worksheet

Date: 7-25-96

Requestor: STEEL

Tank: AN-105
Core: 153
Seg: 21 LH
Auger: _____
Sample ID: _____

Tank: _____
Core: _____
Seg: _____
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10356
Jar/Vial Size: 1 mL
Initial Weight: 27.7 g
Final Weight: 27.7 g
Net Weight: 0 g

Cone#: 56
Final Vol: 9.8 mL
Initial Weight: 7.7 g
Final Weight: 22.1 g
Net Weight: 15.4 g

Sample ID: 5967603971

Appearance/Narrative:

15.4/9.8 = 1.57 g/cc

Jar#: 7
Jar/Vial Size: 1 mL
Initial Weight: 7.7 g
Final Weight: 7.7 g
Net Weight: 0 g

Cone#: _____
Final Vol: _____ mL
Initial Weight: _____ g
Final Weight: _____ g
Net Weight: _____ g

Sample ID: _____

Appearance/Narrative:

Bulk Density Worksheet

Requestor: STW

Date: 7-28-86

Tank: AN-105
Core: 153
Seg: 22 LH
Auger: _____
Sample ID: _____

Tank: AN-105
Core: 153
Seg: 22 LH
Auger: _____
Sample ID: _____

Start Time: _____
End Time: _____
Homogenization Time (Min.): _____

Jar#: 10408
Jar/Vial Size: 4 mL
Initial Weight: 5.4 g
Final Weight: 7.7 g
Net Weight: 2.3 g

Cone#: 08
Final Vol: 10.5 mL
Initial Weight: 7.7 g
Final Weight: 24.3 g
Net Weight: 16.6 g

Sample ID: 586T003973

Appearance/Remarks:

Bulk Density = 16.6/10.5 = 1.58 g/cc

Jar#: 10407
Jar/Vial Size: 1 mL
Initial Weight: 2.7 g
Final Weight: 7.7 g
Net Weight: 5.0 g

Cone#: 07
Final Vol: 10.7 mL
Initial Weight: 7.7 g
Final Weight: 24.8 g
Net Weight: 17.1 g

Sample ID: 586T003972

Appearance/Remarks:

Bulk Density = 17.1/10.7 = 1.60 g/cc

HNF-SD-WM-DF-199, REV. 1

Jul 28 1986 5:02PM

WHC 2225 LAB ROOM 2F BACKSIDE

No. 3438 P. 12/12

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

INORGANIC ANALYSES

HNH-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

THIS PAGE WAS INTENTIONALLY LEFT BLANK

LABCORE Data Entry Template for Worklist# 12274

Analyst: RJM Instrument: DSC0 1 Book # 12N148
Method: LA-514-113 Rev/Mod C-1 HNF-SD-WM-DP-199, REV. 1
Worklist Comment: Please run AN-105 DSC's under N2. bdy

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-01	LIQUID	28.45	36.9	N/A	Joules/g
96000832	AN-105	2 SAMPLE	996T004039	0	DSC-01	LIQUID	N/A	0		Joules/g
96000832	AN-105	3 DUP	996T004039	0	DSC-01	LIQUID	0	0	N/A	Joules/g
96000945	AN-105	4 SAMPLE	996T004305	0	DSC-01	LIQUID	N/A	0		Joules/g
96000945	AN-105	5 DUP	996T004305	0	DSC-01	LIQUID	0	0	N/A	Joules/g

Final page for worklist # 12274

RJM
Analyst Signature

8/26/96
Date

John M. Mc
Analyst Signature

8/29/96
Date

Verified/Validated by
Blandina
Valenzuela 9-3-96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 227 TO 310

DSC STD 12N14-B

4.510 mg

Rate: 10.0 °C/min

File: 00051.004

DSC METTLER

26-Aug-96

Ident: 0.0

222-S Laboratory

6X03

mw °C

Integration
Delta H 121 mJ
26.9 J/g
Peak 158.6 °C
-8.5 mW

120.

140.

160.

180.

°C

Blandina Valenzuela for PD Mayers 9-3-96

556T004039 N2

17.140 mg

Rate: 10.0 °C/min

File: 00059.001

DSC METTLER

27-Aug-96

Ident: 0.0 222-S Laboratory

Exo

20. mW

WHC-SD-WM-OP-199, REV. 1
HNF-SD-WM-OP-199, REV. 1

100.

200.

300.

400.

°C

S96T004039 DUP N2

14.360 mW

Rate: 10.0 °C/min

File: 00061.001 DSC METTLER 27-Aug-96
Ident: 0.0 222-S Laboratory

exo

20. mW

WHC-SD-WM-DP-199, REV. 1
HNF-SD-WM-DP-199, REV. 1

100. 200. 300. 400. °C

S95T004308 N2

14.350 mg

Rate: 10.0 °C/min

File: 00053.001 DSC METTLER 26-Aug-95
Ident: 0.0 222-S Laboratory

ENDO

LABCORE Data Entry Template for Worklist# 12325

Analyst: SHE Instrument: DSC0 1 Book # 820 94 8-28-96
12N14B
Method: LA-514-113 Rev/Mod _____
Worklist Comment: Please run AN-105 DSCs under N2. bdy HNF-SD-WM-DP-199, REV. 1

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-01	SOLID	<u>28.45</u>	<u>28.4*</u>	N/A	Joules/g
96000632	AN-105	2 SAMPLE	S96T004254	0	DSC-01	SOLID	N/A	<u>43.7</u>		Joules/g
96000632	AN-105	3 DUP	S96T004254	0	DSC-01	SOLID	<u>43.7</u>	<u>31.7</u>	N/A	Joules/g *
96000632	AN-105	4 SAMPLE	S96T004259	0	DSC-01	SOLID	N/A	<u>60.4</u>		Joules/g
96000632	AN-105	5 DUP	S96T004259	0	DSC-01	SOLID	<u>60.4</u>	<u>54.1</u>	N/A	Joules/g

Final page for worklist # 12325

Lucie M. Dutton 8-28-96
Analyst Signature Date

Lucie M. Dutton 8-29-96
Analyst Signature Date

* QC FAILURE

Data Entry Comments:

S96T004254 results are the sum of two ^{exotherms} ~~weight loss steps~~
S96T004259 results for the actual are the sum of three exotherms and
the duplicate result is the sum of two exotherms.

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 312 TO 314

DSC STD 12N14B

4.510 mg

Rate: 10.0 °C/min

File: 00063.001

DSC METTLER

28-Aug-98

Ident: 0.0

222-S Laboratory

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

S96T004254 N2
26.741 mg

Rate: 10.0 °C/min

File: 00069.001 DSC METTLER 29-AUG-96
Ident: 0.0 222-S Laboratory

ΔX D B

50. mW

100. 200. 300. 400. °C

WHC-SD-WM-OP-199, REV. 1

HNF-SD-WM-OP-199, REV. 1

S95T004254 DUP N2

20.888 mg

Rate: 10.0 °C/min

File: 00071.001

Ident: 0.0

DSC METTLER

29-Aug-96

222-S Laboratory

x0.2

HNF-SD-WM-OP-199, REV. 1

WHC-SD-WM-OP-199, REV. 1

S96T004259 N2

32.715 mg

Rate: 10.0 °C/min

File: 00055.001

DSC METTLER

28-Aug-96

Ident: 0.0 222-S Laboratory

XOXO

Integration
Delta H 577 mJ
Delta H 17.6 J/g
Peak 240.5°C
3.0 mW

Integration
Delta H 914 mJ
Delta H 27.9 J/g
Peak 312.5°C
3.8 mW

Integration
Delta H 488 mJ
Delta H 14.9 J/g
Peak 414.3°C
1.5 mW

50.0 mW

Integration
Delta H 26136 mJ
Delta H 798.9 J/g
Peak 123.3°C
-60.1 mW

HNF-SD-WM-DP-199, REV. 1

VHC-SD-WM-DP-199, REV. 1

°C

400.

300.

200.

100.

S96T004259 DUP N2

18.578 mg

Rate: 10.0 °C/min

File: 00057.001 DSC METTLER

28-Aug-96

Ident: 0.0 222-S Laboratory

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

LABCORE Data Entry Template for Worklist# 12278

Analyst: RM Instrument: TGA0 1 Book # 92N8AMethod: LA-560-112 Rev/Mod C-1 HNF-SD-WM-DP-199, REV. 1

Worklist Comment: Please run AN-105 TGAs under N2. hdy

GROUP	PROJECT	% TYPE	SAMPLES	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-01	LIQUID	59.2 58.4	58.44	N/A	%
96000632	AN-105	2 SAMPLE	596T004039	0	TGA-01	LIQUID	N/A	53.57	N/A	%
96000632	AN-105	3 DUP	596T004039	0	TGA-01	LIQUID	53.57	46.67	N/A	%
96000945	AN-105	4 SAMPLE	596T004308	0	TGA-01	LIQUID	50.38	50.23	N/A	%
96000945	AN-105	5 DUP	596T004308	0	TGA-01	LIQUID	N/A	53.57	N/A	%
96000945	AN-105						53.57	46.67	N/A	%

Final page for worklist # 12278

Analyst Signature: RM Date: 8/29/96Analyst Signature: Jack Mark Date: 8/29/96Verified/Validated by
Blandina
Valenzuela 9.3.96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Completed Worklist Report for Worklist# 12278

Analyst: rdm

Instrument: TGA01

Book# 82NEA

Method: 1A-560-112 Rev/Mod B-1

Worklist Comment: Please run AN-105 TGAs under N2. bdv

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0	TGA-01	LIGDED	5.02401	58.44	58.718 % Recovery	
2	SAMPLE	5567004028	0	TGA-01	LIGDED	N/A	58.57	%	
3	SUP	5567004039	0	TGA-01	LIGDED	54.77	46.67	13.767 RPD	
4	SAMPLE	5567004100	0	TGA-01	LIGDED	N/A	58.28	%	
5	SUP	5567004108	0	TGA-01	LIGDED	58.20	58.23	0.258 RPD	

Final page for worklist# 12278

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

SIGNATURE BELOW REPRESENTS ANALYST. REPHOTOGRAPH THAT
COMPLETED/ENTERED THE CALIBRATION/ANALYSIS ON PAGES 319 TO 323.

TGA STD 82NB-A
14.926 mg

Rate: 10.0 °C/min

File: 00052.001 TG METTLER 26-Aug-96
Ident: 0.0 222-S Laboratory

Step Analysis
Height -8.72 mg
-58.44 %
Resid. 6.20 mg
41.56 %
Dpeak 84.2 °C

5. mg

50.

100.

150.

200.

°C

Blarina Valeriana for RD Mums 4-3-96

S96T004039 NZ

45.507 mg

Rate: 10.0 °C/min

File: 00660.001 TS METTLER 27-Aug-96

Ident: 0.0

222-S Laboratory

5. mg

S95T004039 DUP N2
20.533 mg

Rate: 10.0 °C/min

File: 00062.001 T6 METTLER 27-AUG-98
Ident: 0.0 222-S Laboratory

Step Analysis
Height -8.31 mg
-40.48 %
Resid. 12.22 mg
59.52 %
Dpeak 85.0 °C

Step Analysis
Height -1.27 mg
-6.19 %
Resid. 10.91 mg
53.15 %

S961004308 N2
17.623 mg

Rate: 10.0 °C/min

File: 00054.001 TG METTLER 26-Aug-96
Ident: 0.0 222-S Laboratory

5. mg

S96T004308DUP N2

19.519 mg

Rate: 10.0 °C/min

File: 00058.001

TS METTLER 27-AUG-96
Ident: 0.0 222-S Laboratory

LABCORE Data Entry Template for Worklist# 12334

Analyst: SNF Instrument: TGA0 1 Book # 82A28A

Method: LA-560-112 Rev/Mod B-1

HNF-SD-WM-DP-199, REV. 1

Worklist Comment: Please run AN-105 TGAs under N2. bdy

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-01	SOLID	59.2	59.08	N/A	%
96000832	AN-105	2 SAMPLE	S96T004254	0	TGA-01	SOLID	N/A	46.19		%
96000832	AN-105	3 DUP	S96T004254	0	TGA-01	SOLID	46.19	45.29	N/A	%
96000832	AN-105	4 SAMPLE	S96T004259	0	TGA-01	SOLID	N/A	42.84		%
96000832	AN-105	5 DUP	S96T004259	0	TGA-01	SOLID	42.84	42.61	N/A	%

Final page for worklist # 12334

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S96T004254 results are the sum of two weight loss steps

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 325 TO 324.

WHC-SD-WM-DP-199, REV. 0
WHC-SD-WM-DP-199, REV. 1

TGA STD 82N8A
19.162 mg

Rate: 10.0 °C/min

File: 00084.001
Ident: 0.0

TG METTLER
232-S Laboratory

28-AUG-96

Step Analysis

Height-11.32 mg

-59.08 %

Resid. 7.84 mg

40.92 %

Dpeak 86.7 °C

5.0 mg

HNF-SD-WM-DP-199, REV. 1

150.

100.

50.

Lucy M. Walker 8/28/96

HNF-SD-WM-DP-199, REV. 1

File: 00070.001 TG METTLER 29-AUG-98
Ident: 0.0 222-S Laboratory596T004254 N2
15.610 mg Rate: 10.0 °C/min

Step Analysis
Height -6.14 mg
-39.36 %
Resid. 9.47 mg
60.64 %

Step Analysis
Height -1.07 mg
-6.83 %
Resid. 8.40 mg
53.81 %

S96T004254 DUP N2
 24.373 mg
 Rate: 10.0 °C/min
 File: 00072.001
 Ident: 0.0
 T6
 METTLER
 29-AUG-96
 222-S Laboratory

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

S96T004259 N2
13.611 mg

Rate: 10.0 °C/min

File: 00086.001 TG METTLER 28-Aug-96
Ident: 0.0 222-S Laboratory

Step Analysis

Height -5.80 mg

-42.61 %

Resid. 7.81 mg

57.39 %

Dpeak 67.0 °C

2. mg

WHC-SD-WM-OP-199, REV. 1
HNF-SD-WM-OP-199, REV. 1

°C

400.

300.

200.

100.

S96T004259 DUP N2

16.708 mg

Rate: 10.0 °C/min

File: 00068.001

TG METTLER 28-Aug-98

Ident: 0.0 223-S Laboratory

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

Step Analysis
Height -7.16 mg
-42.87 %
Resid. 9.54 mg
57.13 %
Dpeak 81.0 °C

LABCORE Data Entry Template for Worklist#

11273

Analyst: RJ Instrument: BA001 Book # 133N16-A

Method: LA-510-112 Rev/Mod C.3

Worklist Comment: AN-105 SPG. RCJ

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LICUID	1.3125	1.378	N/A	Sp.G.
96000832	AN-105	2 SAMPLE	596T003732	0	SPG-01	LICUID	N/A	1.418		Sp.G.
96000832	AN-105	3 DUP	596T003732	0	SPG-01	LICUID	1.418	1.403	N/A	Sp.G.
96000832	AN-105	4 SAMPLE	596T003739	0	SPG-01	LICUID	N/A	1.408		Sp.G.
96000832	AN-105	5 DUP	596T003739	0	SPG-01	LICUID	1.408	1.410	N/A	Sp.G.
96000832	AN-105	6 SAMPLE	596T003740	0	SPG-01	LICUID	N/A	1.406		Sp.G.
96000832	AN-105	7 DUP	596T003740	0	SPG-01	LICUID	1.406	1.410	N/A	Sp.G.
96000832	AN-105	8 SAMPLE	596T003741	0	SPG-01	LICUID	N/A	1.394		Sp.G.
96000832	AN-105	9 DUP	596T003741	0	SPG-01	LICUID	1.394	1.401	N/A	Sp.G.
96000832	AN-105	10 SAMPLE	596T003742	0	SPG-01	LICUID	N/A	1.416		Sp.G.
96000832	AN-105	11 DUP	596T003742	0	SPG-01	LICUID	1.416	1.406	N/A	Sp.G.
96000832	AN-105	12 SAMPLE	596T003743	0	SPG-01	LICUID	N/A	1.399		Sp.G.
96000832	AN-105	13 DUP	596T003743	0	SPG-01	LICUID	1.399	1.408	N/A	Sp.G.

Final page for worklist #

11273

RaeAnn Green 9-8-96
Analyst Signature Date

9-8-96
Analyst Signature Date

Approved RJ Schuchman
9/10/96

Data Entry Comments:

Units shown for QC (SPG & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

INSTRUMENT
CODEANALYSIS
TIMEANALYSIS
DATE

SPECIFIC GRAVITY: LA-510-112 (C-3)

WORKLIST
#ANALYST
SIGNATURE

11273

R. Anderson 09-08-96

11:55

☒ SAMPLE
☐ STANDARD☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # = 133N06A

STD # = 996T003741

CALIB 250.281

STD # = 250.28 ml

TARE WEIGHT (g)	1.3601
GROSS WEIGHT (g)	1.4061
VOL. # SOLUTION (ml)	3.554

TARE WEIGHT (g)	1.3601
GROSS WEIGHT (g)	1.4061
VOL. # SOLUTION (ml)	3.554

☒ SAMPLE
☐ STANDARD☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # = 996T003732

STD # = 996T003742

TARE WEIGHT (g)	1.2457
GROSS WEIGHT (g)	1.4025
VOL. # SOLUTION (ml)	3.554

TARE WEIGHT (g)	1.4424
GROSS WEIGHT (g)	1.4742
VOL. # SOLUTION (ml)	3.554

☒ SAMPLE
☐ STANDARD☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # = 996T003739

STD # = 996T003743

TARE WEIGHT (g)	1.3495
GROSS WEIGHT (g)	1.4025
VOL. # SOLUTION (ml)	3.554

TARE WEIGHT (g)	1.2780
GROSS WEIGHT (g)	1.4222
VOL. # SOLUTION (ml)	3.554

☒ SAMPLE
☐ STANDARD☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # = 996T003740

STD # =

TARE WEIGHT (g)	1.4235
GROSS WEIGHT (g)	1.4742
VOL. # SOLUTION (ml)	3.554

TARE WEIGHT (g)	
GROSS WEIGHT (g)	
VOL. # SOLUTION (ml)	

WHO-SD-WM-OP-199, REV. 1

HNF-SD-WM-OP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STANDARD	STANDARD
STANDARD	Gross Weight (W2)	1.7081	1.6870
Work List	Tare Weight (W1)	1.3607	1.3426
11273	Weight of Solution (W2-W1)	0.3454	0.3444
Test Code	Volume of Solution μ L	250.2800	250.2800
SPG-01	Specific Gravity	1.3801	1.3761
Matrix	Specific Gravity (Average)	1.3781	
LIQUID			
Sample #			
133N16A			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity Average =	1.378	
11:55 AM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7405	
Work List	Tare Weight (W1)	1.3857	
11273	Weight of Solution (W2-W1)	0.3548	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4176	NA
Matrix			
LIQUID			
Sample #			
S96T003732			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.418	
11:55 AM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev. 1.1	Page 1 of 1

WHC-SD-WM-DP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7109	
Work List	Tare Weight (W1)	1.3697	
11273	Weight of Solution (W2-W1)	0.3612	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4032	NA
Matrix			
LIQUID			
Sample #			
S96T003732			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L}/\text{mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g}/\text{mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.403	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7020	
Work List	Tare Weight (W1)	1.3495	
11273	Weight of Solution (W2-W1)	0.3525	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4084	NA
Matrix			
LIQUID			
Sample #			
996T003739			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.408	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510-112.1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7485	
Work List	Tare Weight (W1)	1.3955	
11273	Weight of Solution (W2-W1)	0.353	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4104	NA
Matrix			
LIQUID			
Sample #			
S96T003739			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.410	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7753	
Work List	Tare Weight (W1)	1.4235	
11273	Weight of Solution (W2-W1)	0.3518	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4056	NA
Matrix			
LIQUID			
Sample #			
S96T003740			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.406	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev. 1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7484	
Work List	Tare Weight (W1)	1.3955	
11273	Weight of Solution (W2-W1)	0.3529	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4100	NA
Matrix			
LIQUID			
Sample #			
S96T003740			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.410	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7350	
Work List	Tare Weight (W1)	1.3861	
11273	Weight of Solution (W2-W1)	0.3489	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3940	NA
Matrix			
LIQUID			
Sample #			
S96T003741			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	v RESULT v		
Time	Specific Gravity =	1.394	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7359	
Work List	Tare Weight (W1)	1.3852	
11273	Weight of Solution (W2-W1)	0.3507	0
Test Code	Volume of Solution μ L	280.2800	
SPG-01	Specific Gravity	1.4012	NA
Matrix			
LIQUID			
Sample #			
S96T003741			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.401	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7999	
Work List	Tare Weight (W1)	1.4454	
11273	Weight of Solution (W2-W1)	0.3545	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4164	NA
Matrix			
LIQUID			
Sample #			
S96T003742			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.416	
11:55 AM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7427	
Work List	Tare Weight (W1)	1.3909	
11273	Weight of Solution (W2-W1)	0.3518	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4056	NA
Matrix			
LIQUID			
Sample #			
S96T003742			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.406	
11:55 AM			

Date Entry by: <i>D.H. Hunter</i>	Date: 09/08/96
Approved by: <i>Red Schoeden</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7282	
Work List	Tare Weight (W1)	1.3780	
11273	Weight of Solution (W2-W1)	0.3502	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3892	NA
Matrix			
LIQUID			
Sample #			
S96T003743			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.389	
11:55 AM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7338	
Work List	Tare Weight (W1)	1.3887	
11273	Weight of Solution (W2-W1)	0.3523	0
Test Code	Volume of Solution μL	250.2800	
SPG-01	Specific Gravity	1.4076	NA
Matrix			
LIQUID			
Sample #			
S96T003743			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	✓ RESULT ✓		
Time	Specific Gravity =	1.408	
11:55 AM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev. 1.1	Page 1 of 1

LABCORE Completed Worklist Report for Worklist# 11275

Analyst: adp

Instrument: BA001

Book# 133A116-A

Method: LA-SD-112 Rev/Mod C-3

Worklist Comment: AN-105 SPG. RCJ

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
3 STD	0	SPG-01	LAQ010	1.35752	1.391	85.533 % Recovery	
5 SAMPLE	896T003744 0	SPG-01	LAQ010	N/A	1.437	1.00e-003 Sp. G.	
3 STD	896T003744 0	SPG-01	LAQ010	1.437	1.432	0.343 STD	
4 SAMPLE	896T003745 0	SPG-01	LAQ010	N/A	1.347	1.00e-003 Sp. G.	
5 STD	896T003745 0	SPG-01	LAQ010	1.347	1.353	0.444 STD	
6 SAMPLE	896T003746 0	SPG-01	LAQ010	N/A	1.405	1.00e-003 Sp. G.	
7 STD	896T003746 0	SPG-01	LAQ010	1.405	1.400	0.576 STD	
8 SAMPLE	896T003747 0	SPG-01	LAQ010	N/A	1.267	1.00e-003 Sp. G.	
9 STD	896T003747 0	SPG-01	LAQ010	1.367	1.362	0.072 STD	

Final page for worklist# 11275

Analyst Signature

Date

Analyst Signature

Date

verified in Labcore

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist#

11275

Analyst: ADP Instrument: BA001 Book # 133016-A

Method: LA-510-112 Rev/Mod C-3

Worklist Comment: AN-105 SPG, RCJ

GROUP	PROJECT	S TYPE	SAMPLER	N A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID	1.39785	1.397	N/A	Sp.G. 100% R
96000832	AN-105	2 SAMPLE	5967003744	0	SPG-01	LIQUID	N/A	1.437		Sp.G.
96000832	AN-105	3 DUP	5967003744	0	SPG-01	LIQUID	1.437	1.432	N/A	Sp.G.
96000832	AN-105	4 SAMPLE	5967003745	0	SPG-01	LIQUID	N/A	1.347		Sp.G.
96000832	AN-105	5 DUP	5967003745	0	SPG-01	LIQUID	1.347	1.353	N/A	Sp.G.
96000832	AN-105	6 SAMPLE	5967003746	0	SPG-01	LIQUID	N/A	1.408		Sp.G.
96000832	AN-105	7 DUP	5967003746	0	SPG-01	LIQUID	1.408	1.410	N/A	Sp.G.
96000832	AN-105	8 SAMPLE	5967003747	0	SPG-01	LIQUID	N/A	1.367		Sp.G.
96000832	AN-105	9 DUP	5967003747	0	SPG-01	LIQUID	1.367	1.368	N/A	Sp.G.

Final page for worklist #

11275

Anthony Purnetto 8/12/96
Analyst Signature Date

Susan Bue 8/13/96
Analyst Signature Date

included in Labcore
(print completed worklist)

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit. S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

INSTRUMENT
CODEANALYSIS
TIMEANALYSIS
DATE

SPECIFIC GRAVITY: LA-610-112 (C-3)

WORKLIST

ANALYST
SIGNATURE

8-12-96

☐ SAMPLE
☒ DUPLICATE

SAMPLE # = 34D

STD # = 133N16-1

TARE WEIGHT (g) 1.3398
GROSS WEIGHT (g) 1.3793
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3398
1.3793
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003744

STD # =

TARE WEIGHT (g) 1.3369
GROSS WEIGHT (g) 1.3538
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3369
1.3538
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003744

STD # =

TARE WEIGHT (g) 1.3292
GROSS WEIGHT (g) 1.3236
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3292
1.3236
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003745

STD # =

TARE WEIGHT (g) 1.3970
GROSS WEIGHT (g) 1.3918
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3970
1.3918
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003745

STD # =

TARE WEIGHT (g) 1.4105
GROSS WEIGHT (g) 1.5300
VOL. & SOLUTION (mL) 100.1REPLICATE
1.4105
1.5300
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003746

STD # =

TARE WEIGHT (g) 1.3694
GROSS WEIGHT (g) 1.5084
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3694
1.5084
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003746

STD # =

TARE WEIGHT (g) 1.3344
GROSS WEIGHT (g) 1.4786
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3344
1.4786
100.1☐ SAMPLE
☒ DUPLICATE

SAMPLE # = S96T003747

STD # =

TARE WEIGHT (g) 1.3973
GROSS WEIGHT (g) 1.5392
VOL. & SOLUTION (mL) 100.1REPLICATE
1.3973
1.5392
100.1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STANDARD	STANDARD
STANDARD	Gross Weight (W2)	1.4792	1.5094
Work List	Tare Weight (W1)	1.3398	1.3702
11275	Weight of Solution (W2-W1)	0.1394	0.1392
Test Code	Volume of Solution μL	100.1200	100.1200
SPG-01	Specific Gravity	1.3923	1.3903
Matrix	Specific Gravity (Average)	1.3913	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	v RESULT v		
Time	Specific Gravity Average =	1.391	
05:30 PM			

Date Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96
Form 510112L1 Rev. 1.1		

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.5308	
Work List	Tare Weight (W1)	1.3869	
11275	Weight of Solution (W2-W1)	0.1438	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4373	NA
Matrix			
LIQUID			
Sample #			
S96T003744			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
GA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.437	
05:30 PM			

Date Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.5228	
Work List	Tare Weight (W1)	1.3792	
11275	Weight of Solution (W2-W1)	0.1434	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4323	NA
Matrix			
LIQUID			
Sample #			
S96T003744			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.432	
05:30 PM			

Date Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.5319	
Work List	Tare Weight (W1)	1.3978	
11276	Weight of Solution (W2-W1)	0.1349	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3474	NA
Matrix			
LIQUID			
Sample #			
996T003745			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADF	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	v RESULT v		
Time	Specific Gravity =	1.347	
05:30 PM			

Data Entry by:

Date:

08/13/96

Approved by:

Date:

08/13/96

Form 510112L1 Rev. 1.1

Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.5460	
Work List	Tare Weight (W1)	1.4105	
11275	Weight of Solution (W2-W1)	0.1355	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3634	NA
Matrix			
LIQUID			
Sample #			
S96T003745			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/13/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.363	
05:30 PM			

Data Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96
Form 510112L1 Rev. 1.1		

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.5034	
Work List	Tare Weight (W1)	1.3624	
11275	Weight of Solution (W2-W1)	0.141	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4083	NA
Matrix			
LIQUID			
Sample #			
S96T003746			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.408	
05:30 PM			

Data Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.4798	
Work List	Tare Weight (W1)	1.3394	
11275	Weight of Solution (W2-W1)	0.1402	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4003	NA
Matrix			
LIQUID			
Sample #			
S96T003746			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	∇ RESULT ∇		
Time	Specific Gravity =	1.400	
05:30 PM			

Data Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.5343	
Work List	Tare Weight (W1)	1.3973	
11275	Weight of Solution (W2-W1)	0.1369	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3874	NA
Matrix			
LIQUID			
Sample #			
S96T003747			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	v RESULT v		
Time	Specific Gravity =	1.357	
05:30 PM			

Date Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.5399	
Work List	Tare Weight (W1)	1.4029	
11275	Weight of Solution (W2-W1)	0.137	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3684	NA
Matrix			
LIQUID			
Sample #			
S96T003747			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
08/12/96	v RESULT v		
Time	Specific Gravity =	1.368	
05:30 PM			

Date Entry by:	Date:	08/13/96
Approved by:	Date:	08/13/96

LABCORE Data Entry Template for Worklist#

11276

Analyst: RM Instrument: BA001 Book # 133N16-A

Method: LA-510-112 Rev/Mod C-3

Worklist Comment: AN-105 SPG. RCJ

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	L10U10	<u>1.39765</u>	<u>1.380</u>	<u>N/A</u>	Sp.G.
96000832	AN-105	2 SAMPLE	9961003748	0	SPG-01	L10U10	<u>N/A</u>	<u>1.382</u>		Sp.G.
96000832	AN-105	3 DUP	9961003748	0	SPG-01	L10U10	<u>1.382</u>	<u>1.382</u>	<u>N/A</u>	Sp.G.
96000832	AN-105	4 SAMPLE	9961003749	0	SPG-01	L10U10	<u>N/A</u>	<u>1.392</u>		Sp.G.
96000832	AN-105	5 DUP	9961003749	0	SPG-01	L10U10	<u>1.392</u>	<u>1.383</u>	<u>N/A</u>	Sp.G.
96000832	AN-105	6 SAMPLE	9961003750	0	SPG-01	L10U10	<u>N/A</u>	<u>1.376</u>		Sp.G.
96000832	AN-105	7 DUP	9961003750	0	SPG-01	L10U10	<u>1.376</u>	<u>1.388</u>	<u>N/A</u>	Sp.G.
96000832	AN-105	8 SAMPLE	9961003751	0	SPG-01	L10U10	<u>N/A</u>	<u>1.400</u>		Sp.G.
96000832	AN-105	9 DUP	9961003751	0	SPG-01	L10U10	<u>1.400</u>	<u>1.395</u>	<u>N/A</u>	Sp.G.

Final page for worklist #

11276

Railinn Green 09-08-96
Analyst Signature Date

[Signature] 9-8-96
Analyst Signature Date

Approved Robert W. Schwach
9/10/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

INSTRUMENT
CODEANALYSIS
TIMEANALYSIS
DATE

SPECIFIC GRAVITY: LA-910-112 (C-3)

WORKLIST
#ANALYST
SIGNATURE

11276

Ree Ann Green

9-08-90

16:00

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

CALIB 250.39 mL

SAMPLE # = 133 N16-A STD # =

REPLICATE
1.9139
1.9241
1.7281
2.0117

TARE WEIGHT (g) 1.9139
GROSS WEIGHT (g) 2.2000
VOL. of SOLUTION (mL) 0.2861☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 96T000748 STD # =

REPLICATE
1.2915
1.7310
1.5450

TARE WEIGHT (g) 1.2915
GROSS WEIGHT (g) 1.7310
VOL. of SOLUTION (mL) 0.4450☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 96T000314 STD # =

REPLICATE
1.5076
1.7101
1.5435

TARE WEIGHT (g) 1.5076
GROSS WEIGHT (g) 1.7101
VOL. of SOLUTION (mL) 0.2012☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 96T0003150 STD # =

REPLICATE
1.4000
1.7100
1.5410

TARE WEIGHT (g) 1.4000
GROSS WEIGHT (g) 1.7100
VOL. of SOLUTION (mL) 0.3100☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 96T0003751 STD # =

REPLICATE
1.4512
1.7310
1.5435

TARE WEIGHT (g) 1.4512
GROSS WEIGHT (g) 1.7310
VOL. of SOLUTION (mL) 0.2861☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # =

REPLICATE

TARE WEIGHT (g) _____
GROSS WEIGHT (g) _____
VOL. of SOLUTION (mL) _____☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # =

REPLICATE

TARE WEIGHT (g) _____
GROSS WEIGHT (g) _____
VOL. of SOLUTION (mL) _____☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # =

REPLICATE

TARE WEIGHT (g) _____
GROSS WEIGHT (g) _____
VOL. of SOLUTION (mL) _____

WHC-SD-WM-OP-199, REV. 1

HNF-SD-WM-OP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STANDARD	STANDARD
STANDARD	Gross Weight (W2)	2.2600	1.7288
Work List	Tare Weight (W1)	1.9139	1.3841
11276	Weight of Solution (W2-W1)	0.3461	0.3447
Test Code	Volume of Solution μ L	250.2800	250.2800
SPG-01	Specific Gravity	1.3829	1.3773
Matrix	Specific Gravity (Average)	1.3801	
LIQUID			
Sample #			
133N16A			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	ν RESULT ν		
Time	Specific Gravity Average =	1.380	
04:00 PM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510112L1 Rev 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7375	
Work List	Tare Weight (W1)	1.3915	
11275	Weight of Solution (W2-W1)	0.346	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3825	NA
Matrix			
LIQUID			
Sample #			
S96T003748			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAAG	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.382	
04:00 PM			

Date Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96
Form 510-112.1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7359	
Work List	Tare Weight (W1)	1.3901	
11276	Weight of Solution (W2-W1)	0.3458	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3817	NA
Matrix			
LIQUID			
Sample #			
S96T003748			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAAG	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
09/08/96	$\sqrt{\text{RESULT}}$		
Time	Specific Gravity =	1.382	
04:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7161	
Work List	Tare Weight (W1)	1.2676	
11276	Weight of Solution (W2-W1)	0.3485	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3924	NA
Matrix			
LIQUID			
Sample #			
SRST003749			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.392	
04:00 PM			

Data Entry by: *[Signature]* Date: 09/08/96
 Approved by: *[Signature]* Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7069	
Work List	Tare Weight (W1)	1.3607	
11276	Weight of Solution (W2-W1)	0.3462	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3833	NA
Matrix			
LIQUID			
Sample #			
S96T003749			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	v RESULT v		
Time	Specific Gravity =	1.383	
04:00 PM			

Data Entry by: *[Signature]* Date: 09/08/96
 Approved by: *[Signature]* Date: 9/10/96
 Form 510112L Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.7465	
Work List	Tare Weight (W1)	1.4020	
11276	Weight of Solution (W2-W1)	0.3445	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3765	NA
Matrix			
LIQUID			
Sample #			
S96T003750			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	✓ RESULT ✓		
Time	Specific Gravity =	1.376	
04:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7480	
Work List	Tare Weight (W1)	1.3926	
11276	Weight of Solution (W2-W1)	0.3474	0
Test Code	Volume of Solution μ L	260.2800	
SPG-01	Specific Gravity	1.3880	NA
Matrix			
LIQUID			
Sample #			
S961003760			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	v RESULT v		
Time	Specific Gravity =	1.388	
04:00 PM			

Data Entry by: Date: 09/08/96
 Approved by: Date: 9/10/96
 Form 510112L1 Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.8016	
Work List	Tare Weight (W1)	1.4612	
11276	Weight of Solution (W2-W1)	0.3504	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.4000	NA
Matrix			
LIQUID			
Sample #			
S96T003751			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
09/08/96	✓ RESULT ✓		
Time	Specific Gravity =	1.400	
04:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 09/08/96
Approved by: <i>[Signature]</i>	Date: 9/10/96

Form 510/112.1 Rev 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.7260	
Work List	Tare Weight (W1)	1.3769	
11276	Weight of Solution (W2-W1)	0.3491	0
Test Code	Volume of Solution μ L	250.2800	
SPG-01	Specific Gravity	1.3948	NA
Matrix			
LIQUID			
Sample #			
S96T003751			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RAG	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
09/08/96	▼ RESULT ▼		
Time	Specific Gravity =	1.395	
04:00 PM			

Data Entry by: *[Signature]* Date: 09/08/96
 Approved by: *[Signature]* Date: 9/10/96
 Form 510112L1 Rev 1.1 Page 1 of 1

LABCORE Completed Worklist Report for Worklist# 11666

Analyst: dph

Instrument: BA001

Book# _____

Method: LA570.11Z Rev/Mod C.3

Worklist Comment: AN-105 FOR SPG

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	SPG-01	LIQUID	1.339	1.338	99.242 % Recovery	
2 SAMPLE	8567063574 0	SPG-01	LIQUID	N/A	.562	1.00e-002 Sp. G.	
3 STD	8567063574 0	SPG-01	LIQUID	.252	.257	0.309 STD	
4 SAMPLE	8567063575 0	SPG-01	LIQUID	N/A	1.433	1.00e-003 Sp. G.	
5 STD	8567063575 0	SPG-01	LIQUID	1.433	1.424	0.626 STD	
6 SAMPLE	8567063576 0	SPG-01	LIQUID	N/A	1.676	1.00e-002 Sp. G.	
7 STD	8567063576 0	SPG-01	LIQUID	1.676	1.442	2.330 STD	
8 SAMPLE	8567063577 0	SPG-01	LIQUID	N/A	1.462	1.00e-002 Sp. G.	
9 STD	8567063577 0	SPG-01	LIQUID	1.462	1.637	2.423 STD	
10 SAMPLE	8567063578 0	SPG-01	LIQUID	N/A	2.472	1.00e-002 Sp. G.	
11 STD	8567063578 0	SPG-01	LIQUID	2.472	2.478	0.487 STD	
12 SAMPLE	8567063579 0	SPG-01	LIQUID	N/A	1.454	1.00e-002 Sp. G.	
13 STD	8567063579 0	SPG-01	LIQUID	1.454	1.445	1.385 STD	
14 SAMPLE	8567063580 0	SPG-01	LIQUID	N/A	1.412	1.00e-003 Sp. G.	
15 STD	8567063580 0	SPG-01	LIQUID	1.412	1.408	0.284 STD	

Final page for worklist# 11666

Analyst Signature

Date

Analyst Signature

Date

Rui Schwelb 10/22/96
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 11666

Analyst: DPB Instrument: BA001 Book# 133A/16-AMethod: LA-510-112 Rev/Mod C-3

Worklist Comment: AN-105 FOR SPG

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				SPG-01	LIQUID		
2	SAMPLE	S96T001974 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
3	DUP	S96T001974 0			SPG-01	LIQUID		
4	SAMPLE	S96T001975 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
5	DUP	S96T001975 0			SPG-01	LIQUID		
6	SAMPLE	S96T001976 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
7	DUP	S96T001976 0			SPG-01	LIQUID		
8	SAMPLE	S96T001977 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
9	DUP	S96T001977 0			SPG-01	LIQUID		
10	SAMPLE	S96T001978 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
11	DUP	S96T001978 0			SPG-01	LIQUID		
12	SAMPLE	S96T001979 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
13	DUP	S96T001979 0			SPG-01	LIQUID		
14	SAMPLE	S96T001980 0			SPG-01	LIQUID	96000945	AN-105
		Analytes Requested: SPG-01						
15	DUP	S96T001980 0			SPG-01	LIQUID		

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 11666

S Type Sample# R A Test Matrix Group# Project

Final page for worklist # 11666

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

371

SPECIFIC GRAVITY: LA-510-112 (C-3)

WORKLIST

ANALYST

INITIALS

ANALYSIS

DATE

ANALYSIS

TIME

INSTRUMENT

CODE

11666

DPB

10/20/96

1600

☒ SAMPLE
☐ STANDARD
☐ DUPLICATE

SAMPLE # =

STD # =

133816-A

 TARE WEIGHT (g) 1.8612
 GROSS WEIGHT (g) 2.0689
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.8469
2.0571
149.84 mL

 TARE WEIGHT (g) 1.8396
 GROSS WEIGHT (g) 2.0582
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.8343
2.0478
149.84 mL

SAMPLE # =

STD # =

5967003977

☒ SAMPLE
☐ STANDARD
☐ DUPLICATE

SAMPLE # =

STD # =

5967003974

 TARE WEIGHT (g) 1.8403
 GROSS WEIGHT (g) 2.0675
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.8627
2.0713
149.84 mL

 TARE WEIGHT (g) 1.8195
 GROSS WEIGHT (g) 2.0401
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.7926
2.0293
149.84 mL

SAMPLE # =

STD # =

5967003978

☒ SAMPLE
☐ STANDARD
☐ DUPLICATE

SAMPLE # =

STD # =

5967003975

 TARE WEIGHT (g) 1.8553
 GROSS WEIGHT (g) 2.0701
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.8267
2.1201
149.84 mL

 TARE WEIGHT (g) 1.8494
 GROSS WEIGHT (g) 2.0684
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
1.8277
2.0383
149.84 mL

SAMPLE # =

STD # =

5967003979

☒ SAMPLE
☐ STANDARD
☐ DUPLICATE

SAMPLE # =

STD # =

5967003976

 TARE WEIGHT (g) 1.8682
 GROSS WEIGHT (g) 2.0838
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
2.0875
2.0785
149.84 mL

 TARE WEIGHT (g) 1.8783
 GROSS WEIGHT (g) 2.0843
 VOL. of SOLUTION (mL) 149.84 mL

REPLICATE
2.0873
2.0838
149.84 mL

SAMPLE # =

STD # =

5967003980

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STANDARD	STANDARD
STANDARD	Gross Weight (W2)	2.0884	2.0571
Work List	Tare Weight (W1)	1.8612	1.8489
11666	Weight of Solution (W2-W1)	0.2072	0.2082
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.3823	1.3890
Matrix	Specific Gravity (Average)	1.3857	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
10/20/96	v RESULT v		
Time	Specific Gravity Average =	1.386	
05:00 PM			

SPG Value = 1.39705

Data Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/22/96
Form 510112L1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type	SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9875 2.0113
Work List	Tare Weight (W1)	1.8483 1.8634
11666	Weight of Solution (W2-W1)	0.1472 0.1479
Test Code	Volume of Solution μ L	149.8900 149.8900
SPG-01	Specific Gravity	0.9821 0.9867
Matrix	Specific Gravity (Average)	0.9844
LIQUID		
Sample #		
596T003974		
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution	
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton	
Analyst		
DPB	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$	
Date		
10/20/96	$\checkmark$ RESULT $\checkmark$	
Time	Specific Gravity Average =	0.984
05:00 PM		

Data Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/22/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Duplicate *Sub* *11/27/96*

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0701	2.1001
Work List	Tare Weight (W1)	1.8883	1.8887
11666	Weight of Solution (W2-W1)	0.2148	0.2134
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4231	1.4237
Matrix	Specific Gravity (Average)	1.4284	
LIQUID			
Sample #			
S96T003576			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
10/20/96	∇ RESULT ∇		
Time	Specific Gravity Average =	1.428	
05:00 PM			

Date Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/22/96
Form 510-112L1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Duplicate max 10/21/96

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0896	2.0725
Work List	Tare Weight (W1)	1.8683	1.8564
11666	Weight of Solution (W2-W1)	0.2212	0.2161
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4757	1.4417
Matrix	Specific Gravity (Average)	1.4587	
LIQUID			
Sample #			
S96T003976			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
10/20/96	v RESULT v		
Time	Specific Gravity Average =	1.459	
06:00 PM			

Data Entry by: *[Signature]* Date: 10/21/96
 Approved by: *[Signature]* Date: 10/22/96
 Form 510112L1 Rev. 1.1 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Duplicate

aws
10/20/96

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0588	2.0978
Work List	Tare Weight (W1)	1.8396	1.8839
11666	Weight of Solution (W2-W1)	0.2192	0.2139
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4624	1.4270
Matrix	Specific Gravity (Average)	1.4447	
LIQUID			
Sample #			
S96T003977			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
10/20/96	v RESULT v		
Time	Specific Gravity Average =	1.445	
05:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/20/96
Form 510112L1 Rev. 1.1	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)
Duplicate
per 10/27/96

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0901	2.1242
Work List	Tare Weight (W1)	1.8686	1.9026
11666	Weight of Solution (W2-W1)	0.2206	0.2216
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4717	1.4764
Matrix	Specific Gravity (Average)	1.4751	
LIQUID			
Sample #			
S96T003878			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPE	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
10/20/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity Average =	1.475	
05:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/22/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0688	2.0383
Work List	Tare Weight (W1)	1.8484	1.8217
11666	Weight of Solution (W2-W1)	0.2195	0.2166
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4544	1.4461
Matrix	Specific Gravity (Average)	1.4547	
LIQUID			
Sample #			
S96T003979			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
10/20/96	v RESULT v		
Time	Specific Gravity Average =	1.455	
05:00 PM			

Data Entry by: <i>[Signature]</i>	Date: 10/21/96
Approved by: <i>[Signature]</i>	Date: 10/22/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0899	2.0388
Work List	Tare Weight (W1)	1.8783	1.8278
11686	Weight of Solution (W2-W1)	0.2116	0.211
Test Code	Volume of Solution μ L	149.8900	149.8900
SPG-01	Specific Gravity	1.4117	1.4077
Matrix	Specific Gravity (Average)	1.4097	
LIQUID			
Sample #			
S96T003980			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
DPB	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
10/20/96	v RESULT v		
Time	Specific Gravity Average =	1.410	
06:00 PM			

Data Entry by: <i>John J. Hannon</i>	Date: 10/21/96
Approved by: <i>Ray Schwach</i>	Date: 10/22/96

LABCORE Completed Worklist Report for Worklist# 12398

Analyst: cal

Instrument: BA001

Book#

Method: 1A510.112 Rev/Mod 0.3

Worklist Comment: SPG-01 FOR AN-105

RTS!

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0	SPG-01	L10000	1.380	1.386	96.986 % Recovery	
2	SAMPLE	89820044300	0	SPG-01	L10070	N/A	1.934	1.00e-003	Sp. G.
3	STD	89820044300	0	SPG-01	L10070	1.634	1.484	9.155 EEP	
4	SAMPLE	89820044300	0	SPG-01	L10070	N/A	1.364	1.00e-002	Sp. G.
5	STD	89820044300	0	SPG-01	L10070	1.364	1.464	7.072 REP	
6	SAMPLE	89820044300	0	SPG-01	L10070	N/A	1.460	1.00e-002	Sp. G.
7	STD	89820044300	0	SPG-01	L10000	1.460	1.495	0.345 REP	
8	SAMPLE	89820044300	0	SPG-01	L10070	N/A	1.444	1.00e-003	Sp. G.
9	STD	89820044300	0	SPG-01	L10070	1.444	1.435	0.622 REP	

Final page for worklist# 12398

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Ralph W. Schuch
Reviewer Signature _____ Date 11/7/96

LABCORE Data Entry Template for Worklist# 12398

Analyst: Gal Instrument: BA001 Book # 133N/A

Method: LA-510-112 Rev/Mod E-3

Worklist Comment: SPG-01 FOR AN-105 RTS!

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LITUID	<u>1.29155</u>	<u>1.356</u>	N/A	Sp.G.
96000945	AN-105	2 SAMPLE	9961004308	0	SPG-01	LITUID	N/A	<u>1.5315</u>	<u>0.01</u>	Sp.G.
96000945	AN-105	3 DUP	9961004308	0	SPG-01	LITUID	<u>1.5315</u>	<u>1.4513</u>	N/A	Sp.G.
96000945	AN-105	4 SAMPLE	9961004309	0	SPG-01	LITUID	N/A	<u>1.3615</u>	<u>0.01</u>	Sp.G.
96000945	AN-105	5 DUP	9961004309	0	SPG-01	LITUID	<u>1.3615</u>	<u>1.4635</u>	N/A	Sp.G.
96000945	AN-105	6 SAMPLE	9961004333	0	SPG-01	LITUID	N/A	<u>1.4512</u>	<u>0.01</u>	Sp.G.
96000945	AN-105	7 DUP	9961004333	0	SPG-01	LITUID	<u>1.4512</u>	<u>1.4451</u>	N/A	Sp.G.
96000945	AN-105	8 SAMPLE	9961004419	0	SPG-01	LITUID	N/A	<u>1.4412</u>	<u>0.01</u>	Sp.G.
96000945	AN-105	9 DUP	9961004419	0	SPG-01	LITUID	<u>1.4412</u>	<u>1.4533</u>	N/A	Sp.G.

Final page for worklist # 12398

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

SPECIFIC GRAVITY: LA-510-112 (C-3)

WORKLIST

ANALYST
INITIALSANALYSIS
DATEANALYSIS
TIMERETIREMENT
CODE

17398

20L

11-06-96

1300-716-49

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) 1.8879
 GROSS WEIGHT (g) 2.00274
 VOL. of SOLUTION (ml) 100.000

REPLICATE
1.8879
2.00274
100.000

0.810696 Duplicate

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) 1.92197
 GROSS WEIGHT (g) 2.03750
 VOL. of SOLUTION (ml) 100.000

REPLICATE
1.92197
2.03750
100.000

0.810696 Duplicate

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) 1.81162
 GROSS WEIGHT (g) 2.00274
 VOL. of SOLUTION (ml) 100.000

REPLICATE
1.81162
2.00274
100.000

0.810696 Duplicate

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) 1.89479
 GROSS WEIGHT (g) 2.00274
 VOL. of SOLUTION (ml) 100.000

REPLICATE
1.89479
2.00274
100.000

0.810696 Duplicate

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) 1.91949
 GROSS WEIGHT (g) 2.00274
 VOL. of SOLUTION (ml) 100.000

REPLICATE
1.91949
2.00274
100.000

0.810696 Duplicate

☒ SAMPLE
☐ STANDARD ☒ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) _____
 GROSS WEIGHT (g) _____
 VOL. of SOLUTION (ml) _____

REPLICATE

☒ SAMPLE
☐ STANDARD ☐ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) _____
 GROSS WEIGHT (g) _____
 VOL. of SOLUTION (ml) _____

REPLICATE

☒ SAMPLE
☐ STANDARD ☐ DUPLICATE

SAMPLE # =

STD # =

 TARE WEIGHT (g) _____
 GROSS WEIGHT (g) _____
 VOL. of SOLUTION (ml) _____

REPLICATE

WHC-SD-WM-DP-199, REV. 11/96 WHC-SD-WM-DP-199, REV. 1

383

al
11-6-96

1.86329 g STD.

1.86539 g STD. Dup.

1.92197 g 08

1.87277 g 08 Dup.

1.87168 g 09

1.98045 g 09 Dup.

1.94477 g 33

1.88576 g 33 Dup.

1.91949 g 19

1.86848 g 19 Dup.

2.00728 g Std.

2.02199 g Std. Dup.

2.07060 g 08

2.01837 g 08 Dup.

2.00621 g 09

2.04698 g 09 Dup.

1.98996 g 33

2.07844 g 33 Dup.

2.06488 g 19

2.00410 g 19 Dup.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STD	REPLICATE
STD	Gross Weight (W2)	2.0043	2.0230
Work List	Tare Weight (W1)	1.8693	1.8854
12398	Weight of Solution (W2-W1)	0.135	0.1366
Test Code	Volume of Solution μ L	100.1200	100.1200
BA001	Specific Gravity	1.3484	1.3644
Matrix	Specific Gravity (Average)	1.3564	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/98	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity Average =	1.356	
04:49 PM			

Data Entry by:	<i>Eve P...</i>	Date:	11/06/98
Approved by:	<i>RW Sch...</i>	Date:	1/14/99

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0756	
Work List	Tare Weight (W1)	1.9220	
12358	Weight of Solution (W2-W1)	0.15363	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.5348	NA
Matrix			
LIQUID			
Sample #			
S96T004308			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
11/05/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.534	
04:48 PM			

Data Entry by:

Date:

11/05/96

Approved by:

Date:

1/14/97

Form 510112L1 Rev. 1.1

Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0184	
Work List	Tare Weight (W1)	1.8728	
12395	Weight of Solution (W2-W1)	0.1456	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.4543	NA
Matrix			
LIQUID			
Sample #			
S95T004308			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = WL of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
11/06/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.454	
04:49 PM			

Data Entry by:	<i>Sue Bee</i>	Date:	11/06/96
Approved by:	<i>RW Scherbin</i>	Date:	1/14/97
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0082	
Work List	Tare Weight (W1)	1.8718	
12398	Weight of Solution (W2-W1)	0.1365	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.3645	NA
Matrix			
LIQUID			
Sample #			
S95T004309			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	v RESULT v		
Time	Specific Gravity =	1.364	
04:49 PM			

Data Entry by:	<i>Ge. Perry</i>	Date:	11/06/96
Approved by:	<i>DW. Schmitt</i>	Date:	1/14/97

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0470	
Work List	Tare Weight (W1)	1.8005	
12395	Weight of Solution (W2-W1)	0.146521	0
Test Code	Volume of Solution μL	100.1200	
BA001	Specific Gravity	1.4635	NA
Matrix			
LIQUID			
Sample #			
S86T004309			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	v RESULT v		
Time	Specific Gravity =	1.463	
04:49 PM			

Data Entry by:

Date: 11/06/96

Approved by:

Date: 1/14/97

Form 510112L1 Rev. 1.1

Page 1 of 1

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9900	
Work List	Tare Weight (W1)	1.8448	
12398	Weight of Solution (W2-W1)	0.14519	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.4502	NA
Matrix			
LIQUID			
Sample #			
S95T004333			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
11/06/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.450	
04:49 PM			

Data Entry by:	<i>Sue Rex</i>	Date:	11/06/96
Approved by:	<i>RW Schroder</i>	Date:	11/14/97
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0304	
Work List	Tare Weight (W1)	1.8858	
12398	Weight of Solution (W2-W1)	0.144671	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.4450	NA
Matrix			
LIQUID			
Sample #			
S96T004333			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
11/06/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.445	
04:49 PM			

Data Entry by:	<i>Gwen Bee</i>	Date:	11/06/96
Approved by:	<i>R.W. Schaefer</i>	Date:	1/14/97

WHC-SD-WM-DP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0641	
Work List	Tare Weight (W1)	1.9195	
12398	Weight of Solution (W2-W1)	0.14459	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.4442	NA
Matrix			
LIQUID			
Sample #			
S98T004419			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPG-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.444	
04:49 PM			

Data Entry by:	<i>Sue Ren</i>	Date:	11/06/96
Approved by:	<i>Pat Schneider</i>	Date:	1/14/97
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0041	
Work List	Tare Weight (W1)	1.8604	
12398	Weight of Solution (W2-W1)	0.1437	0
Test Code	Volume of Solution μ L	100.1200	
BA001	Specific Gravity	1.4353	NA
Matrix			
LIQUID			
Sample #			
S95T004419			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
SPQ-01	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	$\vee$ RESULT $\vee$		
Time	Specific Gravity =	1.435	
04:49 PM			

Data Entry by:	<i>C. R. R.</i>	Date:	11/06/96
Approved by:	<i>S. J. J.</i>	Date:	1/14/96

Form 510112L1 Rev. 1.1

Page 1 of 1

LABCORE Data Entry Template for Worklist# 12399

Analyst: ADP Instrument: BA001 Book # 133N16A

Method: LA-510-112 Rev/Mod C-3

Worklist Comment: SPG-01 FOR AN-105 RTS!

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID	<u>1.391</u>	<u>1.407</u>	<u>N/A</u>	Sp.G.
96000945	AN-105	2 SAMPLE	596T004420	0	SPG-01	LIQUID	<u>N/A</u>	<u>1.388</u>	<u>0.000</u>	Sp.G.
96000945	AN-105	3 DUP	596T004420	0	SPG-01	LIQUID	<u>1.387</u>	<u>1.342</u>	<u>N/A</u>	Sp.G.
96000945	AN-105	4 SAMPLE	596T004421	0	SPG-01	LIQUID	<u>N/A</u>	<u>1.501</u>	<u>0.000</u>	Sp.G.
96000945	AN-105	5 DUP	596T004421	0	SPG-01	LIQUID	<u>1.511</u>	<u>1.502</u>	<u>N/A</u>	Sp.G.
96000945	AN-105	6 SAMPLE	596T004433	0	SPG-01	LIQUID	<u>N/A</u>	<u>1.431</u>	<u>0.000</u>	Sp.G.
96000945	AN-105	7 DUP	596T004433	0	SPG-01	LIQUID	<u>1.421</u>	<u>1.446</u>	<u>N/A</u>	Sp.G.
96000945	AN-105	8 SAMPLE	596T004434	0	SPG-01	LIQUID	<u>N/A</u>	<u>1.459</u>	<u>0.000</u>	Sp.G.
96000945	AN-105	9 DUP	596T004434	0	SPG-01	LIQUID	<u>1.469</u>	<u>1.457</u>	<u>N/A</u>	Sp.G.

Final page for worklist # 12399

Analyst Signature ADP Date 12-3-96

Analyst Signature Richard R. Schwach Date 12/03/96

Approved Richard R. Schwach
12/5/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

SPECIFIC GRAVITY: LA-510-112 (C-3)

WIGRELIST

ANALYST

ANALYSIS

INSTRUMENT

DATE

TIME

CODE

12399

A08

12-3-96

1600

☒ SAMPLE
☐ STANDARD☐ DUPLICATE

SAMPLE # =

STD # =

133N16A

TARE WEIGHT (g)

1.8312

GROSS WEIGHT (g)

1.9718

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8316
1.9718
1.0012

☒ SAMPLE
☐ STANDARD☐ DUPLICATE

SAMPLE # =

STD # =

S96T004420

TARE WEIGHT (g)

1.8272

GROSS WEIGHT (g)

1.9762

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8272
1.9762
1.0012

395

☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # =

STD # =

S96T004420

TARE WEIGHT (g)

1.8241

GROSS WEIGHT (g)

1.9842

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8241
1.9842
1.0012

☒ SAMPLE
☐ STANDARD☐ DUPLICATE

SAMPLE # =

STD # =

S96T004421

TARE WEIGHT (g)

1.8409

GROSS WEIGHT (g)

1.9927

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8409
1.9927
1.0012

☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # =

STD # =

S96T004421

TARE WEIGHT (g)

1.8190

GROSS WEIGHT (g)

1.9805

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8190
1.9805
1.0012

☒ SAMPLE
☐ STANDARD☐ DUPLICATE

SAMPLE # =

STD # =

S96T004433

TARE WEIGHT (g)

1.7076

GROSS WEIGHT (g)

2.0509

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.7076
2.0509
1.0012

☒ SAMPLE
☐ STANDARD☒ DUPLICATE

SAMPLE # =

STD # =

S96T004433

TARE WEIGHT (g)

1.8426

GROSS WEIGHT (g)

1.9862

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8426
1.9862
1.0012

☒ SAMPLE
☐ STANDARD☐ DUPLICATE

SAMPLE # =

STD # =

S96T004434

TARE WEIGHT (g)

1.8067

GROSS WEIGHT (g)

1.9529

VOL. of SOLUTION (mL)

1.0012

REPLICATE
1.8067
1.9529
1.0012

WHC-SD-WM-DP-189, REV. 1

HNF-SD-WM-DP-189, REV. 1

SPECIFIC GRAVITY: LA-510-112 (C-3)

INSTRUMENT
CODEANALYSIS
TIMEANALYSIS
DATEANALYST
SIGNATUREWORKLIST
#

12399

12-3-96

1600

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = 5967004434 STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)1.7862
1.8332
1.0002

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

☒ SAMPLE
STANDARD☒ DUPLICATE

SAMPLE # = STD # =

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

REPLICATE

WHC-SD-WM-OP-199, REV. 1

HNF-DSD-WM-OP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STD	REPLICATE
STD	Gross Weight (W2)	1.9718	2.0008
Work List	Tare Weight (W1)	1.8312	1.8696
12399	Weight of Solution (W2-W1)	0.1406	0.1412
Test Code	Volume of Solution μ L	100.1200	100.1200
SPG-01	Specific Gravity	1.4043	1.4103
Matrix	Specific Gravity (Average)	1.4073	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	v RESULT v		
Time	Specific Gravity Average =	1.407	
04:00 PM			

Data Entry by:	ADP	Date:	12/03/96
Approved by:	BW Schrock	Date:	12/5/96
Form 518112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9762	
Work List	Tare Weight (W1)	1.8372	
12399	Weight of Solution (W2-W1)	0.139	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3883	NA
Matrix			
LIQUID			
Sample #			
S96T004420			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	v RESULT v		
Time	Specific Gravity =	1.388	
04:00 PM			

Date Entry by:	<i>Sue Acc</i>	Date:	12/03/96
Approved by:	<i>R.W. Schreder</i>	Date:	12/5/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.8545	
Work List	Tare Weight (W1)	1.8201	
12399	Weight of Solution (W2-W1)	0.1344	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.3424	NA
Matrix			
LIQUID			
Sample #			
S96T004420 DUP			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	ν RESULT ν		
12/03/96	Specific Gravity =	1.342	
Time			
04:00 PM			

Date Entry by:	<i>S. P. Lee</i>	Date:	12/03/96
Approved by:	<i>B. J. Orveden</i>	Date:	12/14/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9922	
Work List	Tare Weight (W1)	1.8409	
12399	Weight of Solution (W2-W1)	0.1513	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.5112	NA
Matrix			
LIQUID			
Sample #			
S961004421			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	ν RESULT ν		
Time	Specific Gravity =	1.511	
04:00 PM			

Data Entry by:	<i>SWP</i>	Date:	12/03/96
Approved by:	<i>RW Schneider</i>	Date:	12/15/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9695	
Work List	Tare Weight (W1)	1.8190	
12399	Weight of Solution (W2-W1)	0.1505	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.5032	NA
Matrix			
LIQUID			
Sample #			
S95T004421 DUP			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
12/03/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.503	
04:00 PM			

Data Entry by:	<i>S. R. Bap</i>	Date:	12/03/96
Approved by:	<i>R. W. Schreder</i>	Date:	12/5/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0509	
Work List	Tare Weight (W1)	1.9076	
12399	Weight of Solution (W2-W1)	0.1433	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4313	NA
Matrix			
LIQUID			
Sample #			
S96T004433			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.431	
04:08 PM			

Date Entry by:	<i>Sue Rye</i>	Date:	12/03/96
Approved by:	<i>R.W. Schroeder</i>	Date:	12/5/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9868	
Work List	Tare Weight (W1)	1.8428	
12399	Weight of Solution (W2-W1)	0.1442	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4403	NA
Matrix			
LIQUID			
Sample #			
S95T004433 DUP			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.440	
04:00 PM			

Data Entry by:	<i>Sue Roe</i>	Date:	12/03/96
Approved by:	<i>D.W. Schwartz</i>	Date:	12/14/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9628	
Work List	Tare Weight (W1)	1.8067	
12399	Weight of Solution (W2-W1)	0.1461	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4692	NA
Matrix			
LIQUID			
Sample #			
S96T004434			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
12/03/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.469	
04:00 PM			

Data Entry by:	<i>SUP Rev</i>	Date:	12/03/96
Approved by:	<i>RW Schuch</i>	Date:	12/5/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9322	
Work List	Tare Weight (W1)	1.7862	
12399	Weight of Solution (W2-W1)	0.146	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4583	NA
Matrix			
LIQUID			
Sample #			
996T004434 DUP			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
ADP	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date			
12/03/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.458	
04:00 PM			

Data Entry by:	<i>Sue Bee</i>	Date:	12/03/96
Approved by:	<i>Bruce Schrock</i>	Date:	12/5/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

LABCORE Completed Worklist Report for Worklist# 14536

Analyst: cal Instrument: BA001 Book#

Method: 18510.112 Rev/Mod C-2

Worklist Comment: SPG-01 FOR (AN-105) S-302 GRAB(244-TX GRAB) RTS!

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	SPG-01	L1Q01D	1.258	1.280	98.732 % Recovery	
2 SAMPLE	8967005546 0	SPG-01	L1Q01D	N/A	1.432	1.00e-002 Sp.G.	
3 DUP	8967005546 0	SPG-01	L1Q01D	1.422	1.408	0.969 RPD	
4 SAMPLE	2967005556 0	SPG-01	L1Q01D	N/A	.987	1.00e-003 Sp.G.	
5 DUP	2967005556 0	SPG-01	L1Q01D	.987	.984	0.996 RPD	
6 SAMPLE	8967005587 0	SPG-01	L1Q01D	N/A	.979	1.00e-003 Sp.G.	
7 DUP	8967005587 0	SPG-01	L1Q01D	.972	.960	0.204 RPD	
8 SAMPLE	8967005598 0	SPG-01	L1Q01D	N/A	.958	1.00e-002 Sp.G.	
9 DUP	2967005598 0	SPG-01	L1Q01D	.945	.926	0.161 RPD	
10 SAMPLE	8967005643 0	SPG-01	L1Q01D	N/A	1.053	1.00e-003 Sp.G.	
11 DUP	8967005643 0	SPG-01	L1Q01D	1.052	1.062	0.946 RPD	
12 SAMPLE	8967005665 0	SPG-01	L1Q01D	N/A	1.057	1.00e-002 Sp.G.	
13 DUP	8967005665 0	SPG-01	L1Q01D	1.057	1.060	0.282 RPD	
14 SAMPLE	8967005666 0	SPG-01	L1Q01D	N/A	1.062	1.00e-002 Sp.G.	
15 DUP	2967005666 0	SPG-01	L1Q01D	1.043	1.023	2.769 RPD	

Final page for worklist# 14536

Analyst Signature Date

Analyst Signature Date

Reviewer Signature Date 11/7/96

LABCORE Data Entry Template for Worklist# 14536

Analyst: EAL Instrument: BA001 Book # 133WCAMethod: LA-510-112 Rev/Mod C.3

Worklist Comment: SPG-01 FOR (AN-105)(S-302 GRAB)(244-TX GRAB) RTS!

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID	<u>1.3775</u>	<u>1.3880</u>	N/A	Sp.G.
96000832	AN-105	2 SAMPLE	5961005546	0	SPG-01	LIQUID	N/A	<u>1.4222</u>		Sp.G.
96000832	AN-105	3 DUP	5961005546	0	SPG-01	LIQUID	<u>1.4222</u>	<u>1.4077</u>	N/A	Sp.G.
96001418	S-302 GRAB	4 SAMPLE	5961005596	0	SPG-01	LIQUID	N/A	<u>1.9872</u>		Sp.G.
96001418	S-302 GRAB	5 DUP	5961005596	0	SPG-01	LIQUID	<u>1.9872</u>	<u>1.9839</u>	N/A	Sp.G.
96001418	S-302 GRAB	6 SAMPLE	5961005597	0	SPG-01	LIQUID	N/A	<u>1.9885</u>		Sp.G.
96001418	S-302 GRAB	7 DUP	5961005597	0	SPG-01	LIQUID	<u>1.9885</u>	<u>1.9797</u>	N/A	Sp.G.
96001418	S-302 GRAB	8 SAMPLE	5961005598	0	SPG-01	LIQUID	N/A	<u>1.9887</u>		Sp.G.
96001418	S-302 GRAB	9 DUP	5961005598	0	SPG-01	LIQUID	<u>1.9887</u>	<u>1.9877</u>	N/A	Sp.G.
96001429	244-TX GRAB	10 SAMPLE	5961005663	0	SPG-01	LIQUID	N/A	<u>1.0518</u>		Sp.G.
96001429	244-TX GRAB	11 DUP	5961005663	0	SPG-01	LIQUID	<u>1.0518</u>	<u>1.0616</u>	N/A	Sp.G.
96001429	244-TX GRAB	12 SAMPLE	5961005665	0	SPG-01	LIQUID	N/A	<u>1.0574</u>		Sp.G.
96001429	244-TX GRAB	13 DUP	5961005665	0	SPG-01	LIQUID	<u>1.0574</u>	<u>1.0605</u>	N/A	Sp.G.
96001429	244-TX GRAB	14 SAMPLE	5961005666	0	SPG-01	LIQUID	N/A	<u>1.0616</u>		Sp.G.
96001429	244-TX GRAB	15 DUP	5961005666	0	SPG-01	LIQUID	<u>1.0616</u>	<u>1.0327</u>	N/A	Sp.G.

Final page for worklist # 14536

A. Spindel 11/6/96Sue B. B. 11/6/96

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Site Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 14536

GROUP	PROJECT	S TYPE	SAMPLE	R A	TEST	WATERY	ACTUAL	FOUND	DL	UNIT
Analyst Signature		Date		Analyst Signature		Date				

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

SPECIFIC GRAVITY ANALYSIS

ac-
 1146-96
 GROSS Wt.
 910 1.97528 g
 STD 2.82378 g
 46 1.96888 g
 46Dup 2.85351 g
 96 2.84518 g
 96Dup 1.95433 g
 97 2.88339 g
 97Dup 1.95532 g
 98 1.99422 g
 98Dup 2.81117 g
 63 2.85725 g
 63Dup 2.84889 g
 65 1.95272 g
 65Dup 1.95831 g
 66 1.95753 g
 66Dup 1.95312 g

ac-1146-96
 TARE Wt.
 STD 1.83748 g
 STD 1.88888 g
 46 1.81851 g
 46Dup 1.91257 g
 96 1.84925 g
 96Dup 1.89584 g
 97 1.90542 g
 97Dup 1.90123 g
 98 1.89523 g
 98Dup 1.91288 g
 63 1.95154 g
 63Dup 1.93468 g
 65 1.84685 g
 65Dup 1.88213 g
 66 1.85126 g
 66Dup 1.84973 g

SPECIFIC GRAVITY: LA-510-112 (C-3)

WORKSHEET
#

ANALYST

ANALYSIS
DATEANALYSIS
TIMEINSTRUMENT
CODE

14530

KOL

11-6-96

0930-1215

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # =

STD # = 153N16-A

AL 11.6%

DILUENT

1.84140
1.84140
1.84140
1.84140
1.84140

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.84925
1.84925
1.84925
1.84925
1.84925

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.84023
1.84023
1.84023
1.84023
1.84023

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.84083
1.84083
1.84083
1.84083
1.84083

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.83816
1.83816
1.83816
1.83816
1.83816

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.90594
1.90594
1.90594
1.90594
1.90594

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.95194
1.95194
1.95194
1.95194
1.95194

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

☒ SAMPLE
☒ STANDARD☐ DUPLICATE

SAMPLE # = 946T005940

STD # =

AL 11.6%

DILUENT

1.87126
1.87126
1.87126
1.87126
1.87126

TARE WEIGHT (g)
GROSS WEIGHT (g)
VOL. of SOLUTION (mL)

WHC-SD-WM-DP-199, REV. 1

HNF-SD-WM-DP-199, REV. 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		STD	REPLICATE
STD	Gross Weight (W2)	1.9758	2.0237
Work List	Tare Weight (W1)	1.8374	1.8858
14836	Weight of Solution (W2-W1)	0.1384	0.1379
Test Code	Volume of Solution μL	100.1200	100.1200
SPG-01	Specific Gravity	1.3823	1.3773
Matrix	Specific Gravity (Average)	1.3788	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	v RESULT v		
Time	Specific Gravity Average =	1.380	
12:15 PM			

Data Entry by:	Sue Bee	Date:	11/06/96
Approved by:	RW. Schroeder	Date:	11/8/96
Form 510112LJ Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.8609	
Work List	Tare Weight (W1)	1.8185	
14836	Weight of Solution (W2-W1)	0.14239	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4222	NA
Matrix			
LIQUID			
Sample #			
S95T006646			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = [(W2-W1) * 1000 μ L/mL] / [Vol. of Solution μ L * 1.000 g/mL]		
Date			
11/06/96	v RESULT v		
Time	Specific Gravity =	1.422	
12:15 PM			

Data Entry by:	<i>Steve Bess</i>	Date:	11/06/96
Approved by:	<i>R.W. Schneider</i>	Date:	11/8/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (C-3)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0535	
Work List	Tare Weight (W1)	1.9126	
14536	Weight of Solution (W2-W1)	0.14094	0
Test Code	Volume of Solution μ L	100.1200	
SPG-01	Specific Gravity	1.4077	NA
Matrix			
LIQUID			
Sample #			
S95T005545			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
EAL	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L}/\text{mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
11/06/96	$\checkmark$ RESULT $\checkmark$		
Time	Specific Gravity =	1.408	
12:15 PM			

Data Entry by:	<i>Sue Bee</i>	Date:	11/06/96
Approved by:	<i>R.W. Schuch</i>	Date:	11/8/96
Form 510112L1 Rev. 1.1		Page 1 of 1	

LABCORE Completed Worklist Report for Worklist# 11911

Analyst: raw

Instrument: PH01

Book# 79118

Method: LA-211-102 Rev/Mod C-0

Worklist Comment: AN-107 (PRV) AND AN-105 OH, PER R. SCHROEDER. RCJ

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLK	0	OH-01	LIVORG	1	<42		ug/gb
2 STD	0	OH-01	LIVORG	1.566e4	1.62e4	97.15% Recovery	
3 SAMPLE	S96D000164 0	OH-01	LIVORG	N/A	128	123.000	ug/gb
4 STD	S54D000164 0	OH-01	LIVORG	<128	<250		RPD
5 SAMPLE	S96D000174 0	OH-01	LIVORG	N/A	42	41.000	ug/gb
6 STD	S54D000174 0	OH-01	LIVORG	<42	<42		RPD

Comments Section:

Comments for sample# S96D000164 and test OH-01

Sample contains no free OH due to demonstrated buffer effect
raws 8/14/96

Final page for worklist# 11911

see attached
Analyst Signature Date

R. W. Schroeder 8/14/96
Analyst Signature Date

in lab core
Reviewer Signature Date
R. W. Schroeder 8/14/96

LABCORE Data Entry Template for Worklist#

11911

Analyst: raw Instrument: PH01 Book # 79N8

Method: LA-211-102 Rev/Mod C-0

Worklist Comment: AN-107 (PRV) AND AN-105 OH, PER R. SCHROEDER. RCI

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK			OH-01	LIQUID	<u>1</u>	<u><42</u>	N/A	ug/ml
		2 STD			OH-01	LIQUID	<u>1.664</u>	<u>1.664</u>	N/A	ug/ml
96001099	AN-107 (PRV)	3 SAMPLE	9960000164	0	OH-01	LIQUID	<u>N/A</u>	<u><125</u>	<u>125</u>	ug/ml
96001099	AN-107 (PRV)	4 DUP	9960000164	0	OH-01	LIQUID	<u><125</u>	<u><250</u>	N/A	ug/ml
96000945	AN-105	5 SAMPLE	9961003974	0	OH-01	LIQUID	<u>5.341</u>	<u>5.341</u>	<u>4.172</u>	ug/ml
96000945	AN-105	6 DUP	9961003974	0	OH-01	LIQUID	<u>5.341</u>	<u>5.341</u>	<u>4.172</u>	ug/ml

Final page for worklist #

11911

Analyst Signature R. Schroeder Date 8/13/96

Analyst Signature _____ Date _____

Data Entry Comments:

Sample 9961000164 contains no free OH
due to demonstrated buffer effect R. Schroeder 8/14/96

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

WHC-SD-WM-DP-199, REV. 1

calibration data
 date 96-08-13
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.cel. 22.5 °C
 slope(ref) 1.007
 U(tax) -12.9 mV
 electr. input 1
 =====

date 96-08-13 time 08:53
 GET pH 12 # 156
 Id.#1 01
 Id.#2 .2005
 pH(init) 11.58
 V/ml pH
 EP1 .238 9.79
 EP2 .327 6.00
 manual stop
 =====

44 79AB 1.64E4 mV

date 96-08-13 time 08:53
 GET pH 12 # 156
 .10ml/div dH=1/div
 start V .000 ml

date 96-08-13 time 08:54
 GET pH 12 # 157
 Id.#1 00
 Id.#2 .2005
 pH(init) 4.70
 V/ml pH
 EP1 .000 3.90
 manual stop
 =====

date 96-08-13 time 08:55
 GET pH 12 # 157
 .10ml/div dH=1/div
 start V .000 ml

date 96-08-13 time 09:17
 GET pH 12 # 159
 Id.#1 164 576000164
 Id.#2 .2005
 pH(init) 10.55
 manual stop
 =====

1000ml

date 96-08-13 time 09:24
 GET pH 12 # 160
 Id.#1 164 576000164
 Id.#2 .2005
 pH(init) 10.31
 stop V reached
 =====

date 96-08-13 time 09:26
 GET pH 12 # 160
 .10ml/div dH=1/div
 start V .000 ml

dup 500ml
 date 96-08-13 time 09:33
 GET pH 12 # 162
 Id.#1 164 576000164 Dup
 Id.#2 .2005
 pH(init) 9.32
 stop V reached
 =====

date 96-08-13 time 09:33
 GET pH 12 # 162
 .10ml/div dH=1/div
 start V .000 ml

WHO-SD-WM-DP-199, REV. 1

pH 10.87
pH 10.87
pH 9.93
pH 9.93

Sample
3.0 ml

date 96-08-13 time 09:51
GET pH 12 # 164
Id.#1 3974 S96T003974
Id.#2 .2005
pH(initial) 9.95
V/ml pH
EP1 .016 9.36
EP2 .047 6.74
manual stop
=====

date 96-08-13 time 09:52
GET pH 12 # 164
.10ml/div dppH=1/div
start V .000 ml

Dup
3.0 ml

pH 10.83
pH 10.84
pH 9.88
pH 9.88

date 96-08-13 time 09:58
GET pH 12 # 165
Id.#1 3974 S96T003974 Dup
Id.#2 .2005
pH(initial) 9.90
V/ml pH
EP1 .016 9.31
EP2 .037 7.63
EP3 .058 5.44
manual stop
=====

date 96-08-13 time 09:58
GET pH 12 # 165
.10ml/div dppH=1/div
start V .000 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		Bik
Type	Sample Size (mL) SS	3.000
Bik	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.008
11911	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	5.35E-04
OH-01	OH in Sample in µg/mL (PPM)	9.09E+00
Matrix		
Liquid		
Sample #	Detection Limit = 125µg / SS * DF	
Blank		
Instrument Code	Detection Limit (µg/mL)	4.17E+01
PH01		
Analyst	OH Molarity = (mL HNO3) * (M HNO3) / Sample Size in mL * Dilution Factor	
RAW		
Date	OH in µg/mL = (OH MOLARITY) * (17.0g/mole) * ((1000000µg/g) / (1000mL/L))	
08/13/96		
Time		
		Bik
Concentration of OH in Sample (Molarity)		5.35E-04
OH in Sample in µg/mL (PPM)		<42

The Result is < Detection Limit

Data Entry by:	RUS	Date:	08/13/96
Approved by:	DW Schrodin	Date:	08/17/97
Form 211102_1 Rev. 1.3		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO ₃ (Molarity)	0.2005
Work List	HNO ₃ Titrant at OH end-point in mL	0.238
11911	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	9.54E-01
OH-01	OH in Sample in µg/mL (PPM)	1.62E+04
Matrix		
Liquid		
Sample #	Detection Limit = 125µg / SS * DF	
Std		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = (mL HNO ₃) * (M HNO ₃) / Sample Size in mL * Dilution Factor	
RAW		
Date	OH in µg/mL = (OH MOLARITY) * (17.0g/mole) * ((1000000µg/g) / (1000mL/L))	
08/13/95		
Time		
	Concentration of OH in Sample (Molarity)	9.54E-01
	OH in Sample in µg/mL (PPM)	1.62E+04

Data Entry by:	BUS	Date:	08/14/95
Approved by:	BW Schwach	Date:	11/1/97
Form 211102_1 Rev. 1.3		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		Sample
Type	Sample Size (mL) SS	3.000
Sample	Concentration of HNO3 (Molarity)	0.2006
Work List	HNO3 Titrant at OH end-point in mL	0.016
11911	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.07E-03
OH-01	OH in Sample in µg/mL (PPM)	1.82E+01

Matrix		
Liquid		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003974		
Instrument Code	Detection Limit (µg/mL)	4.17E+01
PH01		

Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
---------	--	--

RAW		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	

08/13/96		
Time		

	Sample
Concentration of OH in Sample (Molarity)	1.07E-03
OH in Sample in µg/mL (PPM)	<42

The Result is < Detection Limit

Data Entry by:	<i>RW</i>	Date:	08/14/96
Approved by:	<i>RW</i>	Date:	1/7/97
Form 211102_1 Rev. 1.3		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

Type	Sample Size (mL) SS	Dup
Dup	Concentration of HNO3 (Molarity)	3.000
Work List	HNO3 Titrant at OH end-point in mL	0.2005
11911	Dilution Factor DF	0.016
Test Code	Concentration of OH in Sample (Molarity)	1
OH-01	OH in Sample in µg/mL (PPM)	1.07E-03
Matrix		1.82E+01
Liquid		
Sample #	Detection Limit = 125µg / SS * DF	
S96T003974		
Instrument Code	Detection Limit (µg/mL)	4.17E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RAW		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
08/13/96		
Time		
		Dup
	Concentration of OH in Sample (Molarity)	1.07E-03
	OH in Sample in µg/mL (PPM)	<42

The Result is < Detection Limit

Data Entry by:	<i>RWS</i>	Date:	08/14/96
Approved by:	<i>RW Schindler</i>	Date:	1/7/97
Form 211102_1 Rev. 1.3			

Page 1 of 1

LABCORE Completed Worklist Report for Worklist# 12791

Analyst: rwk

Instrument: PH01

Book# 7948

Method: LA-211-102 Rev/Mod C.D.

Worklist Comment: AN-105 OH. RCJ

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	OR-01	LIGAND	1.664e4	1.689e4	100.962 % Recovery	
2 BLANK	0	OR-01	LIGAND	3	2.18e1	23.900 ug/dL	
3 SAMPLE	8987004831	OR-01	LIGAND	N/A	2.46e4	2500.000 ug/dL	
4 TOP	8987004831	OR-01	LIGAND	5.56e4	6.05e4	6.651 STD	

Final page for worklist# 12791

See attached
Analyst Signature _____ Date _____

See attached
Analyst Signature _____ Date _____

verified into Labview

John H. Chubb 09/12/96
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 12791

Analyst: RK Instrument: PH01 Book # 7978

Method: LA-211-102 Rev/Mod C-0

Worklist Comment: AN-105 OH, RCI

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			OH-01	L1QUID	<u>1.664e⁻⁴</u>	<u>1.68e⁻⁴</u>	N/A	ug/ml
		2 BLNK			OH-01	L1QUID	<u>1</u>	<u>2.39e⁻¹</u>	N/A	ug/ml
96080832	AN-105	3 SAMPLE	996T004831	0	OH-01	L1QUID	N/A	<u>2.66e⁻⁴</u>	<u>2.88e⁻⁴</u>	ug/ml
96080832	AN-105	4 DUP	996T004831	0	OH-01	L1QUID	<u>2.66e⁻⁴</u>	<u>6.05e⁻⁴</u>	N/A	ug/ml

Final page for worklist # 12791

Nick Nung 9-10-96
Analyst Signature Date

Suom Bo 9/10/96
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit. S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

OR (AUTO): LA-211-102 (C-0)

WORKSHEET
#

17791

ANALYST
INITIALS

RL

ANALYSIS
DATE

9-10-96

ANALYSIS
TIME

2346

INSTRUMENT
CODE

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # =

SAMPLE SIZE (mL) 2.01

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED 1mL

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # = 7998

SAMPLE SIZE (mL) 0.050mL

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED 1mL

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX = 4.9

SAMPLE STD # = 5487004921

SAMPLE SIZE (mL) 0.050mL

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED 1mL

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # = 5487004921

SAMPLE SIZE (mL) 0.050mL

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED 1mL

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # =

SAMPLE SIZE (mL)

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # =

SAMPLE SIZE (mL)

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # =

SAMPLE SIZE (mL)

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED

* ATTACH PRINT REPORT #2 & #5

☒ SAMPLE STANDARD ☐ DUPLICATE PREP BLANK ☐ SPIN REAGENT BLANK

MATRIX =

SAMPLE STD # =

SAMPLE SIZE (mL)

VOL. H₂O ADDED IN DILUTION (IF NECESSARY)

VOL. DILUTED SAMPLE ANALYZED

* BSC12 VOLUME ADDED

* ATTACH PRINT REPORT #2 & #5

WHO-SD-WM-OP-199, REV. 1

HNF-SD-WM-OP-199, REV. 1

Calibration data
 date 96-09-10
 GET pH 12 # 231
 Id.#1 4331
 Id.#2 2095
 pH 10.46
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 15:48
 GET pH 12 # 231
 Id.#1 4331
 Id.#2 2095
 pH 10.46
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 20:00
 GET pH 12 # 230
 Id.#1 2095
 Id.#2 11.68
 pH 9.79
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 21:31
 GET pH 12 # 230
 Id.#1 2095
 Id.#2 11.68
 pH 9.79
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 23:09
 GET pH 12 # 232
 Id.#1 4331
 Id.#2 2095
 pH 10.42
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 23:02
 GET pH 12 # 232
 Id.#1 4331
 Id.#2 2095
 pH 10.42
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 22:50
 GET pH 12 # 231
 Id.#1 4331
 Id.#2 2095
 pH 10.46
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

date 96-09-10 time 22:03
 GET pH 12 # 231
 Id.#1 4331
 Id.#2 2095
 pH 10.46
 stop V reached
 start V .900 ml
 stop V reached
 start V .900 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

Type		BLNK
BLNK	Sample Size (mL) SS	1.000
	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.007
12791	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.40E-03
OH-01	OH in Sample in µg/mL (PPM)	2.39E+01
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
BLNK		
Instrument Code	Detection Limit (µg/mL)	1.25E+02
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL*Dilution Factor	
RWR		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/10/96		
Time		BLNK
11:45 PM	Concentration of OH in Sample (Molarity)	1.40E-03
	OH in Sample in µg/mL (PPM)	<125

The Result is < Detection Limit

Data Entry by:	Date: 09/11/96
Approved by:	Date: 09/11/96
Form 211102_1 Rev. 1.3 Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.248
12791	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	9.86E-01
OH-01	OH in Sample in µg/mL (PPM)	1.68E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
STD		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/10/96		
Time		STD
11:45 PM	Concentration of OH in Sample (Molarity)	9.86E-01
	OH in Sample in µg/mL (PPM)	1.68E+04

Data Entry by

Approved by:

Form 211102_1 Rev. 1.3

Date: 09/11/96

Date: 09/10/96

Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.831
12791	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.33E+00
OH-01	OH in Sample in µg/mL (PPM)	5.66E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S96T004831		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
08/10/96		
Time		SAM
11:45 PM	Concentration of OH in Sample (Molarity)	3.33E+00
	OH in Sample in µg/mL (PPM)	5.66E+04

Data Entry by	Date: 08/11/96
Approved by	Date: 08/11/96
Form 211102_1 Rev. 1.3	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.887
12791	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.56E+00
OH-01	OH in Sample in µg/mL (PPM)	6.06E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S96T004831		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL * Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*((17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/10/96		
Time		DUP
11:46 PM	Concentration of OH in Sample (Molarity)	3.56E+00
	OH in Sample in µg/mL (PPM)	6.06E+04

Data Entry by:	Date:	09/11/96
Approved by:	Date:	09/11/96

LABCORE Data Entry Template for Worklist# 12794

Analyst: _____ Instrument: PH01 _____ Book # _____

Method: LA-211-102 Rev/Mod _____

Worklist Comment: AN-105 OH, RCJ

GROUP	PROJECT	S. TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK			OH-01	SOLID	1	63 ^{new} 10/1/96		ug/g
		2 BLNK-PREP			OH-01	SOLID	1	21.83 ⁴	N/A	ug/g
		3 STD			OH-01	SOLID	1.66e ⁴	1.62e ⁴	N/A	ug/g
96000832	AN-105	4 SAMPLE	9961004258	0 W	OH-01	SOLID	N/A	2.36e ⁴	1.33e ⁴	ug/g
96000832	AN-105	5 DUP	9961004258	0 W	OH-01	SOLID	2.76e ⁴	2.45e ⁴	N/A	ug/g
96000832	AN-105	6 SAMPLE	9961004253	0 W	OH-01	SOLID	N/A	2.24e ⁴	1.18e ⁴	ug/g
96000832	AN-105	7 DUP	9961004263	0 W	OH-01	SOLID	2.24e ⁴	2.10e ⁴	N/A	ug/g

Final page for worklist # 12794

Analyst Signature _____ Date _____

NE Wright 10/1/96
Analyst Signature _____ Date _____

Approved RW Schach 10/3/96

Data Entry Comments:

Units shown for QC (SPE & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 12794

Analyst: RJ Instrument: PH01 Book # 71V8Method: LA-211-102 Rev/Mod C-0

Worklist Comment: AN-105 OH. RCJ

GROUP	PROJECT	S TYPE	SAMPLE#	N A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK			OH-01	SOLID	<u>1</u>	<u><63</u>	N/A	ug/g
		2 STD			OH-01	SOLID	<u>1.66e4</u>	<u>1.62e4</u>	N/A	ug/g
96000832	AN-105	3 SAMPLE	S96T004258	0 N	OH-01	SOLID	N/A	<u>2.36e4</u>	<u>1.33e4</u>	ug/g
96000832	AN-105	4 DUP	S96T004258	0 N	OH-01	SOLID	<u>2.36e4</u>	<u>2.45e4</u>	N/A	ug/g
96000832	AN-105	5 SAMPLE	S96T004263	0 N	OH-01	SOLID	N/A	<u>2.24e4</u>	<u>1.18e4</u>	ug/g
96000832	AN-105	6 DUP	S96T004263	0 N	OH-01	SOLID	<u>2.24e4</u>	<u>2.16e4</u>	N/A	ug/g

Final page for worklist # 12794

Rae Ann Green 9-30-96
 Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

 Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Alias Code.

OH (AUTOR: LA-241-102 (C-0))

WFO/SLT
#ANALYST
INITIALS

DATE

ANALYSIS
TIMEINSTRUMENT
CODE

12-19-94

M. Brown

09-29-96

19:40

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # = 7418

MATRIX =

SAMPLE SIZE (mL) 0.050 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.08

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # =

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.08

☒ SAMPLE
STANDARD☒ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # =

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.07

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # = 5967004258

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.05

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # = 5967004258

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.10

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # = 5967004263

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.10

☒ SAMPLE
STANDARD☒ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # = 5967004263

MATRIX =

SAMPLE SIZE (mL) 3 mL

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED 1 mL

* ATTACH PRINT REPORT #2 & #5

, 0.10

☒ SAMPLE
STANDARD☐ DUPLICATE
PREP BLANK☐ SPIKE
REAGENT BLANK

SAMPLE STD # =

MATRIX =

SAMPLE SIZE (mL)

VOL. H₂O ADDED TO DILUTION (if necessary)

VOL. DILUTED SAMPLE ANALYZED

* BAC12 VOLUME ADDED

* ATTACH PRINT REPORT #2 & #5

Calibration

calibration data
 date 96-09-30 time 17:39
 pH(S) 1 10.00
 pH(S) 2 7.00
 t-cal. 21.0 °C
 slope(frel) 1.000
 U(tas) -11.1 mV
 electr. impd 1

Reagent Blank

date 96-09-30 time 17:39
 GET pH 12 # 35
 Id.#1 00000
 Id.#2 .2000
 pH(init) 5.00
 V(mV) pH
 EP1 .000 4.45
 stop volt.reached

date 96-09-30 time 17:39
 GET pH 12 # 35
 .10ml/div 200ml/div
 start V .000 ml

STD

S9604258249-308

date 96-09-30 time 17:45
 GET pH 12 # 36
 Id.#1 00000
 Id.#2 .2000
 pH(init) 11.95
 V(mV) pH
 EP1 .230 9.71
 EP2 .302 7.03
 stop volt.reached

date 96-09-30 time 17:46
 GET pH 12 # 36
 .10ml/div 200ml/div
 start V .000 ml

Prep Blank

date 96-09-30 time 22:16
GET pH 12 # 50
Id.#1 00000
Id.#2 .2000
pH(init) 4.57

V/ml pH
EP1 .007 4.07
stop volt. reached

date 96-09-30 time 22:17
GET pH 12 # 50
.10ml/div .20ml/div
start V .000 ml

S96T004258

date 96-09-30 time 19:16
GET pH 12 # 43
Id.#1 00000
Id.#2 .2000
pH(init) 11.96

V/ml pH
EP1 .065 9.65
EP2 .156 7.58
stop volt. reached

date 96-09-30 time 19:11
GET pH 12 # 43
.10ml/div .20ml/div
start V .000 ml

S96T004258Dp

date 96-09-30 time 19:18
 GET pH 12 # 44
 Id.#1 00000
 Id.#2 .2005
 pH(init) 11.84
 Vvol Vvol pH
 EP1 .078 10.12
 EP2 .152 8.44
 EP3 .249 5.96
 stop volt.reached

date 96-09-30 time 19:18
 GET pH 12 # 44
 .10ml/div 200ml/div
 start V .000 ml

S96T004263

date 96-09-30 time 19:24
 GET pH 12 # 45
 Id.#1 4283
 Id.#2 .2005
 pH(init) 11.81
 Vvol Vvol pH
 EP1 .079 10.20
 EP2 .161 7.39
 EP3 .205 6.56
 stop volt.reached

date 96-09-30 time 19:25
 GET pH 12 # 45
 .10ml/div 200ml/div
 start V .000 ml

S96T004263-
Dup

date 96-09-08 time 19:31
GET pH 12 # 46
ld.#1 4263
ld.#2 1005
pH(initial) 11.00
Vref Vref pH
EP1 .068 10.15
EP2 .162 9.17
EP3 .228 6.36
stop volt.reached

date 96-09-08 time 19:32
GET pH 12 # 46
.10ml/div pH=1/div
start V .000 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		PREP-BLANK
Type	Sample Size (mL) SS	2.000
PREP-BLANK	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.007
12794	Dilution Factor DF	212.6
Test Code	Concentration of OH in Sample (Molarity)	1.49E-01
OH-01	OH in Sample in µg/mL (PPM)	2.64E+03

Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
PREP-BLANK		
Instrument Code	Detection Limit (µg/mL)	1.33E+04
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time		
		PREP-BLANK
		Concentration of OH in Sample (Molarity)
		1.49E-01
		OH in Sample in µg/mL (PPM)
		<13288
The Result is < Detection Limit		

Data Entry by: <i>W. Wright</i>	Date: 10/01/96
Approved by: <i>W. Johnson</i>	Date: 10/3/96
Form 211102_1 Rev. 1.3	
Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		BLANK
Type	Sample Size (mL) SS	2.000
BLANK	Concentration of HNO3 (Molarity)	0.2008
Work List	HNO3 Titrant at OH end-point in mL	0.008
12794	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	8.02E-04
OH-01	OH in Sample in µg/mL (PPM)	1.36E+01
Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
BLANK		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL*Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time		BLANK
06:39 PM	Concentration of OH in Sample (Molarity)	8.02E-04
	OH in Sample in µg/mL (PPM)	<63
The Result is < Detection Limit		

Data Entry by: NW Date: 10/01/96
 Approved by: RAU Date: 10/1/96
 Form 211102_1 Rev. 1.3 Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		STANDARD
Type	Sample Size (mL) SS	0.060
STANDARD	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.238
12794	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	9.54E-01
OH-01	OH in Sample in µg/mL (PPM)	1.62E+04
Matrix		
SOLID		
Sample #		
79N8		
Instrument Code		
PH01		
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL*Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time	STANDARD	
05:39 PM	Concentration of OH in Sample (Molarity)	9.54E-01
	OH in Sample in µg/mL (PPM)	1.62E+04

Data Entry by:	NW	Date:	10/01/96
Approved by:	<i>Kevin Schneider</i>	Date:	10/1/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

		Sample
Type	Sample Size (mL) SS	2.000
Sample	Concentration of HNO ₃ (Molarity)	0.2005
Work List	HNO ₃ Titrant at OH end-point in mL	0.065
12794	Dilution Factor DF	212.6
Test Code	Concentration of OH in Sample (Molarity)	1.39E+00
OH-01	OH in Sample in µg/mL (PPM)	2.36E+04
Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
S55T004258		
Instrument Code	Detection Limit (µg/mL)	1.33E+04
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL * Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time		
	Concentration of OH in Sample (Molarity)	1.39E+00
	OH in Sample in µg/mL (PPM)	2.36E+04

8
09/30/96

Data Entry by:	RWS	Date:	10/01/96
Approved by:	RW Schmitt	Date:	10/1/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		Sample
Type	Sample Size (mL) SS	2.000
Sample	Concentration of HNO3 (Molarity)	0.2006
Work List	HNO3 Titrant at OH end-point in mL	0.070
12794	Dilution Factor DF	205.1
Test Code	Concentration of OH in Sample (Molarity)	1.44E+00
OH-01	OH in Sample in µg/mL (PPM)	2.45E+04
Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
S96T004258		
Instrument Code	Detection Limit (µg/mL)	1.28E+04
PH01		
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL * Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
09/30/96		
Time		
		Sample
	Concentration of OH in Sample (Molarity)	1.44E+00
	OH in Sample in µg/mL (PPM)	2.45E+04

8
Ratig/Var

Data Entry by	RA/G	Date:	10/01/96
Approved by	RA/G	Date:	10/11/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		Sample
Type	Sample Size (mL) SS	2.000
Sample	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.070
12794	Dilution Factor DF	188.1
Test Code	Concentration of OH in Sample (Molarity)	1.32E+00
OH-01	OH in Sample in µg/mL (PPM)	2.24E+04
Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
S96T004263		
Instrument Code	Detection Limit (µg/mL)	1.18E+04
PH01		
Analyst	OH Molarity = (mL HNO3)*(M HNO3)/Sample Size in mL * Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time		
		Sample
	Concentration of OH in Sample (Molarity)	1.32E+00
	OH in Sample in µg/mL (PPM)	2.24E+04

RWS 10/1/96

Data Entry by	RWS	Date:	10/01/96
Approved by	RWS	Date:	10/1/96

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (C-0)

OH (AUTO) : LA-211-102 (C-0)		DUPLICATE
Type	Sample Size (mL) SS	2.000
DUPLICATE	Concentration of HNO3 (Molarity)	0.2005
Work List	HNO3 Titrant at OH end-point in mL	0.068
12754	Dilution Factor DF	186.4
Test Code	Concentration of OH in Sample (Molarity)	1.27E+00
OH-01	OH in Sample in µg/mL (PPM)	2.16E+04
Matrix		
SOLID		
Sample #	Detection Limit = 125µg / SS * DF	
S96T004263		
Instrument Code	Detection Limit (µg/mL)	1.17E+04
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RA GREEN		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
09/30/96		
Time		
	Concentration of OH in Sample (Molarity)	1.27E+00
	OH in Sample in µg/mL (PPM)	2.16E+04

8 6005
10/1/96

Data Entry by:	RA GREEN	Date:	10/01/96
Approved by:	M. W. Schach	Date:	10/1/96

LABCORE Data Entry Template for Worklist# 12803

Analyst: WV Instrument: BA001 ✓ Book # IN19-F

Method: LA-265-101 Rev/Mod B-0

Worklist Comment: AN-105 CRVL RCJ

LINE	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
1	BLNK				CRV1-01	L10UIR	<u>1.5</u>	<u><0.220</u>	N/A	UG/ML
2	STD				CRV1-01	L10UIR	<u>5.00</u>	<u>5.24</u>	N/A	UG/ML
96000832	AN-105	3 SAMPLE	996T004831	0	CRV1-01	L10UIR	<u>N/A</u>	<u>1.5462</u>	<u>1.5462</u>	UG/ML
96000832	AN-105	4 DUP	996T004831	0	CRV1-01	L10UIR	<u>1.5462</u>	<u>1.5462</u>	<u>1.5462</u>	UG/ML
96000832	AN-105	5 SPK	996T004831	0	CRV1-01	L10UIR	<u>1.00</u>	<u>1.058</u>	N/A	UG/ML

Final page for worklist # 12803

Dale J. Mason 12-16-96
Analyst Signature Date

H. Onesto 12-23-96
Analyst Signature Date

Data Entry and Validation

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Quantum Data Collection Sheet

WHC-SO-WM-DF-199, REV. 1 HNF-SO-WM-DF-199, REV. 1

Worklist Number: 12803

Date/Time: _____

Spike Enoch Number: 11419 F

Spike Volume: _____

ABS *[Signature]*

[illegible]

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		blink	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	
blink	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 * Vsamp) (MF)	11	10
12803	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
996T004831	Cr +6 µg/mL =	<0.220	ERR

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vim	v RESULT v	blink	DUPLICATE
Date	Cr +6 µg/mL =	<0.220	ERR
12/16/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Data Entry by: <i>[Signature]</i>	Date: 12/19/96
Approved by: <i>[Signature]</i>	Date: 1/6/97
Form 265101LT Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		std	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.381	
std	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 + Vsamp) (MF)	11	10
12803	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.88537	0.88537
Sample #	Slope (b)	0.75285	0.75285
5551004831	Cr +6 µg/mL =	5.28E+00	ERR
Calibration Date			
12/13/95	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP]*100		
sp01	Less Than Value = abs < Low Calibration Point		
Analyst			
vim	v RESULT v	std	DUPLICATE
Date	Cr +6 µg/mL =	5.28E+00	ERR
12/16/95			
Time			

The Duplicate Absorbance is < Detection Limit (the Low Calibration Point

RESULT > Recovery % Difference = N/A

Data Entry by: <i>Carl McLaughlin</i>	Date: 12/19/95
Approved by: <i>John P. Papp</i>	Date: 12/19/95
Form 265101LT Rev 1.2	Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.114	0.113
sample	Sample Volume in mL (Vsamp)	1.00	1.00
Work List	Method Factor (10 * Vsamp) (MF)	11	11
12803	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRV1-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
336T004831	Cr +6 µg/mL =	1.54E+02	1.53E+02
Calibration Date			
12/13/96	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP]*100		
sp01	Less Than Value = abs < Low Calibration Point		
Analyst			
vlm	v RESULT v	sample	DUPLICATE
Date			
12/16/96	Cr +6 µg/mL =	1.54E+02	1.53E+02
Time			

RESULT > Recovery % Difference = 0.92%

Data Entry by: <i>[Signature]</i>	Date: 12/18/96
Approved by: <i>[Signature]</i>	Date: 12/18/96
Form 265101LT Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Sample
Detection Limit

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	0.000
sample	Sample Volume in mL (Vsamp)	1.00	1.00
Work List	Method Factor (10 * Vsamp) (MF)	11	11
12803	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
956T004831	Cr +6 µg/mL =	<22.182	<22.182

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP]*100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vin	v RESULT v	sample	DUPLICATE
Date	Cr +6 µg/mL =	<22.182	<22.182
12/16/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Data Entry By: <i>[Signature]</i>	Date: 12/19/96
Approved by: <i>[Signature]</i>	Date: 11/19/97
Form 265101LT Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE			SPIKE
Type	SAMPLE ABSORBANCE at 540 nm	(SAM abs)	0.114
SPIKE	SAMPLE + SPIKE ABSORBANCE at 540 nm	(S+S abs)	0.46
Work List	SAMPLE VOLUME (Sample) in mL	(Vsam)	1.000
12803	SAMPLE VOLUME (Spiked Sample) in mL	(S+S Vsam)	1.000
Test Code	VOLUME of SPIKE in mL	(VOL Spk)	1.000
CRVI-01	SPIKE DILUTION FACTOR	(SDF)	1
Main	SPIKE VALUE	(µg/mL)	5
Liquid			
Sample #	LOW CALIBRATION POINT		0.021
S951004431	HIGH CALIBRATION POINT		0.779
Calibration Date	Y-INTERCEPT	(a)	0.00537
12/13/96	SLOPE	(b)	0.78285
Instrument Code			
sp01	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b	(MSV)	
Analyst	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b	(MSSV)	
Vol	TOTAL VOLUME for SAMP VIAL = 10 + Vsam	(TVS)	
Date	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk	(TVSS)	
12/16/96	% Spike Rec. = [(MSSV * TVSS) - (MSV * TVS * SSCF)] / (SPIKE VALUE * VOL Spk / SDF)		
Time	Note: SAMPLE SIZE CORRECTION FACTOR (SSCF) = S+S Vsam / Vsam		

RESULT % SPIKE RECOVERY = 108.6%

Data entry by:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 265101X3 Rev. 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

Type	Sample Absorbance at 540 nm (abs)	Blank	DUPLICATE
Blank	Sample Volume in ml. (Vsamp)		
Work List	Method Factor (10 + Vsamp) (MF)		
10000	Dilution Factor (DF)		
Test Code	Digest Dilution Factor (DDF)		
Matrix	Low Calibration Point	0.021	
LIQUID	High Calibration Point	0.779	
Sample #	Y-Intercept (a)	0.00537	
	Slope (b)	0.78286	
Calibration Date	Cr +6 µg/ml. =		

12/13/98 Cr +6 µg/ml. = (abs-a) * MF * DF * DDF / (b * Vsamp)

Instrument Code RPD = [(Cr +6 µg/ml. SAM - Cr +6 µg/ml. DUP) / AVERAGE OF THE SAM + DUP] * 100

Less Than Value = abs < Low Calibration Point

Analyst	RESULT	Blank	DUPLICATE
12/13/98	Cr +6 µg/ml. =		

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT > N/A

Data Entry by: *[Signature]* Date: 12/16/98

Approved by: *[Signature]* Date: 12/17/98

Form 285101LT Rev 1.2 Page 1 of 1

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

	std	DUPLICATE
Type	Sample Absorbance at 840 nm (abs)	0.021
std	Sample Volume in mL (Vsamp)	1.0 mL
Work List	Method Factor (10 * Vsamp) (MF)	
LA-265	Dilution Factor (DF)	
Test Code	Digest Dilution Factor (DDF)	
	Low Calibration Point	0.021
Matrix	High Calibration Point	0.775
LIQUID	Y-Intercept (a)	0.00537
Sample #	Slope (b)	0.78285
	Cr +6 µg/mL =	

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)	
12/13/96	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP]*100	
Instrument Code	Less Than Value = abs < Low Calibration Point	
Analyst	v. RESULT v	
Date	Cr +6 µg/mL =	
LA-265-101		
Time		

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT > N/A

Data Entry by:	Date:	12/16/96
Approved by:	Date:	12/16/96

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.114	0.113
sample	Sample Volume in mL (Vsamp)	100.0 - 100.0 - 100.0	100.0 - 100.0 - 100.0
Work List	Method Factor (10 * Vsamp) (MF)		
	Dilution Factor (DF)		
Test Code	Digest Dilution Factor (DDF)		
	Low Calibration Point	0.021	
Matrix	High Calibration Point	0.778	
LIQUID	y Intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
5-16-96	Cr +6 µg/mL	=	
Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/95	RPD = ((Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP) * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
Analyst	v RESULT v		
12-16-96			
Date			
12-16-96	Cr +6 µg/mL =		
Time			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

N/A

Date Entry by: <i>[Signature]</i>	Date: 12/16/96
Approved by: <i>[Signature]</i>	Date: 12/16/96

Form 26510-1LT Rev. 1.2

Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-D) HEXAVALENT CHROMIUM - SPIKE		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.114
SPIKE 11N (1 - 6)	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.410
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	10.00
1-A-50.5	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	10.00
Test Code	VOLUME of SPIKE in mL (VOL Spk)	1.00
	SPIKE DILUTION FACTOR (SDF)	
Matrix	SPIKE VALUE (µg/mL)	
Sample #	LOW CALIBRATION POINT	0.921
5-11-13-14-15	HIGH CALIBRATION POINT	0.779
Calibration Date	Y-INTERCEPT (a)	0.00537
12/13/95	SLOPE (b)	0.79285
Instrument Code		
	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSV)	
Analyst	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)	
Calculated	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)	
Date	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)	
12-13-95	% Spike Rec. = (MSSV * TVSS) / SPIKE VALUE * VML Spk / SDF	
Time		

RESULT	% SPIKE RECOVERY	ERR
--------	------------------	-----

Date entry by:	<i>[Signature]</i>	Date:	12/13/95
Approved by:	<i>[Signature]</i>	Date:	12/13/95

LABCORE Data Entry Template for Worklist# 12805

Analyst: NLM Instrument: BA(X) ✓ Book # 1119F

Method: LA-265-101 Rev/Mod B-0

Worklist Comment: AN-105 CRVI RCJ

GROUP	PROJECT	S TYPE	SAMPLER	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK			CRVI-01	SOLID	<u>0.23</u>	<u>0.23</u>	N/A	ug/s
		2 STD			CRVI-01	SOLID	<u>5.00</u>	<u>5.28</u>	N/A	ug/s
96000832	AN-105	3 SAMPLE	S96T004258	0 W	CRVI-01	SOLID	<u>1.0</u>	<u>0.230</u>	N/A	ug/s
96000832	AN-105	4 DUP	S96T004258	0 W	CRVI-01	SOLID	<u>1.1822</u>	<u>1.2012</u>	N/A	ug/s
96000832	AN-105	5 SPK	S96T004258	0 W	CRVI-01	SOLID	<u>1.00</u>	<u>1.068</u>	N/A	ug/s
96000832	AN-105	6 SAMPLE	S96T004263	0 W	CRVI-01	SOLID	<u>N/A</u>	<u>1.0702</u>	<u>41.31</u>	ug/s
96000832	AN-105	7 DUP	S96T004263	0 W	CRVI-01	SOLID	<u>1.0702</u>	<u>1.0602</u>	N/A	ug/s
96000832	AN-105	8 SPK	S96T004263	0 W	CRVI-01	SOLID	<u>1.00</u>	<u>1.065</u>	N/A	ug/s

Final page for worklist # 12805

Analyst Signature: Valerie L. Mason Date: 12-16-96

Analyst Signature: JL Mason Date: 12-23-96
 Data Entry + Validation

Data Entry Comments: Run prep Blanks with Batch Also read = 0.001

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		blank prep	DUPLICATE
Type	Sample Absorbance at 840 nm (abs)	0.001	
blank prep	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 * Vsamp) (MF)	11	10
12805	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
99ET004253	Cr +6 µg/mL =	<0.220	ERR

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vim	v RESULT v	blank prep	DUPLICATE
Date	Cr +6 µg/mL =	<0.220	ERR
12/16/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Date Entry by:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 265101L1 Rev. 1.2		Page	1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Sample
Det. Limit

Cr+6 : LA-265-101 (B-0)

SOLID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	0.000
sample	Sample Volume in ml. (Vsamp)	1.000	1.000
Work List	Method Factor (10 * Vsamp) (MF)	11	11
12805	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	8.318	5.368
CRIV-01	Low calibration abs point	0.021	0.021
Matrix	High calibration abs point	0.779	0.779
SOLID	Y-intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
S96T004263	Cr +6 µg/g =	<41.313	<40.928
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
vim	v RESULT v	sample	DUPLICATE
Date	Cr +6 µg/g =	<41.313	<40.928
12/15/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Date Entry by	<i>[Signature]</i>	Date	12/19/96
Approved by	<i>[Signature]</i>	Date	12/19/96
Form 35510152 Rev. 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		blank	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	
blank	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 + Vsamp) (MF)	11	10
12885	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
S95T004254	Cr +6 µg/mL =	<0.220	ERR

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vlm	v RESULT v	blank	DUPLICATE
Date			
12/16/96	Cr +6 µg/mL =	<0.220	ERR
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Data Entry by:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Sample
Det. Limit

Cr+6 : LA-265-101 (B-0) SOLID SAMPLE		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	0.000
sample	Sample Volume in ml. (Vsamp)	1.000	1.000
Work List	Method Factor (10 * Vsamp) (MF)	11	11
12805	Dilution Factor (DF)	1	1
Test Code	Digest grl. (DGL)	4.703	4.876
CRIV-01	Low calibration abs point	0.021	0.021
Matrix	High calibration abs point	0.779	0.779
SOLID	Y-intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
5861004258	Cr +6 µg/g =	<46.698	<45.041
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000ml/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
vim	v RESULT v	sample	DUPLICATE
Date	Cr +6 µg/g =	<46.698	<45.041
12/18/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Date Entry by:	<i>[Signature]</i>	Date:	12/18/96
Approved by:	<i>[Signature]</i>	Date:	12/18/96
Form 26510152 Rev. 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE			SPIKE
Type	SAMPLE ABSORBANCE at 540 nm	(SAM abs)	0.046
SPIKE	SAMPLE + SPIKE ABSORBANCE at 540 nm	(S+S abs)	0.39
Work List	SAMPLE VOLUME (Sample) in mL	(Vsam)	1.000
12805	SAMPLE VOLUME (Spiked Sample) in mL	(S+S Vsam)	1.000
Test Code	VOLUME of SPIKE in mL	(VOL Spk)	1.000
CRVI-01	SPIKE DILUTION FACTOR	(SDF)	1
Matrix	SPIKE VALUE	(µg/mL)	5
solid			
Sample #	LOW CALIBRATION POINT		0.021
596T004263	HIGH CALIBRATION POINT		0.779
Calibration Date	Y-INTERCEPT	(a)	0.00537
12/13/96	SLOPE	(b)	0.78285
Instrument Code			
sp01	$Cr + 6 \text{ Molarity in SAMP VIAL} = (SAM \text{ abs} - a) / b \quad (MSV)$		
Analyst	$Cr + 6 \text{ Molarity in S+S VIAL} = (S+S \text{ abs} - a) / b \quad (MSSV)$		
v/m	TOTAL VOLUME for SAMP VIAL = $10 + V_{sam} \quad (TVS)$		
Date	TOTAL VOLUME for S+S VIAL = $10 + S+S V_{sam} + VOL \text{ Spk} \quad (TVSS)$		
12/17/96	$\% \text{ Spike Rec.} = ((MSSV * TVSS) - (MSV * TVS * SSCF)) / (SPIKE \text{ VALUE} * VOL \text{ Spk} / SDF)$		
Time	Note: SAMPLE SIZE CORRECTION FACTOR (SSCF) = $S+S V_{sam} / V_{sam}$		

RESULT	% SPIKE RECOVERY	=	106.5%
--------	------------------	---	--------

Data entry by:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 265101X3 Rev. 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		std	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.381	
std	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 + Vsamp) (MF)	11	10
12805	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78288	0.78288
S98T004258 and 4263	Cr +6 µg/mL =	5.28E+00	ERR
Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = ((Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP)*100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vlm	v RESULT v	std	DUPLICATE
Date	Cr +6 µg/mL =	5.28E+00	ERR
12/18/96			
Time			

The Duplicate Absorbance is < Detection Limit (the Low Calibration Point

RESULT > Recovery % Difference = N/A

Data Entry by:	<i>[Signature]</i>	Date:	12/18/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 26510-1.1 Rev 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) SOLID SAMPLE		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.045	0.047
sample	Sample Volume in mL (Vsamp)	1.000	1.000
Work List	Method Factor (10 * Vsamp) (MF)	11	11
12805	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	4.703	4.876
CRIV-01	Low calibration abs point	0.021	0.021
Matrix	High calibration abs point	0.779	0.779
SOLID	Y-intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
8961004256	Cr +6 µg/g =	1.18E+02	1.20E+02
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs - a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
vim	v RESULT v	sample	DUPLICATE
Date	Cr +6 µg/g =	1.18E+02	1.20E+02
12/15/96			
Time			

RESULT > Recovery % Difference = 1.31

Date Entry By:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 35510152 Rev. 1.2		Page 1 of 1	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.045
SPIKE	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.39
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	1.000
12805	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	1.000
Test Code	VOLUME of SPIKE in mL (VOL Spk)	1.000
CRVI-01	SPIKE DILUTION FACTOR (SDF)	1
Matrix	SPIKE VALUE (µg/mL)	5
solid		
Sample #	LOW CALIBRATION POINT	0.021
S96T004258	HIGH CALIBRATION POINT	0.779
Calibration Date	Y-INTERCEPT (a)	0.00537
12/13/96	SLOPE (b)	0.78285
Instrument Code		
sp01	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSV)	
Analyst	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)	
vtm	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)	
Date	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)	
12/17/96	% Spike Rec = [(MSSV * TVSS) - (MSV * TVS * SSCF)] / (SPIKE VALUE * VOL Spk / SDF)	
Time	Note: SAMPLE SIZE CORRECTION FACTOR (SSCF) = S+S Vsam / Vsam	

RESULT % SPIKE RECOVERY = 106.8%

Data entry by: *C. Jones* Date: 12/18/96
 Approved by: *J. Jones* Date: 12/17/96
 Form 265101X3 Rev. 1.2 Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) SOLID SAMPLE		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.048	0.044
sample	Sample Volume in mL (Vsamp)	1.000	1.000
Work List	Method Factor (10 + Vsamp) (MF)	11	11
12805	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	5.318	5.366
CRIV-01	Low calibration abs point	0.021	0.021
Matrix	High calibration abs point	0.779	0.779
SOLID	Y intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
5981004253	Cr +6 µg/g *	1.07E+02	1.01E+02
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
vim	v RESULT v	sample	DUPLICATE
Date			
12/16/96	Cr +6 µg/g =	1.07E+02	1.01E+02
Time			

RESULT > Recovery % Difference = 5.98

Date Entry by:	<i>[Signature]</i>	Date:	12/19/96
Approved by:	<i>[Signature]</i>	Date:	12/19/96
Form 26510152 Rev. 1-2		Page 1 of 1	

Cr+6 : LA-265-101 (B-0) SOLID SAMPLE		blank	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	
blank	Sample Volume in mL (Vsamp)	1.0 mL	
Work List	Method Factor (10 + Vsamp) (MF)		
12-605	Dilution Factor (DF)		
Test Code	Digest g/L (DGL)		
	Low calibration abs point	0.021	
Matrix	High calibration abs point	0.779	
SOLID	Y-intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
Blank	Cr +6 µg/g =		
Calibration Data			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
Value µg/g	✓ RESULT ✓	blank	DUPLICATE
Date			
12-16-96	Cr +6 µg/g =		
Time			
13:59			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

Data Entry by:	<i>[Signature]</i>	Date:	12/16/96
Approved by:	<i>[Signature]</i>	Date:	12/16/96
Form 265101S2 Rev. 1.2		Page 1 of 1	

Cr+6 : LA-265-101 (B-0)

SOLID SAMPLE

		std	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.861	
std 11N19-F	Sample Volume in mL (Vsamp)	1.0 mL	
Work List	Method Factor (10 + Vsamp) (MF)		
12805	Dilution Factor (DF)		
Test Code	Digest g/L (DGL)		
	Low calibration abs point	0.021	
Matrix	High calibration abs point	0.779	
SOLID	Y-intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
11N19-F	Cr +6 µg/g =		
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
Analyst	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUPI / AVERAGE OF THE SAM + DUP * 100		
10/10/96	Less Than Value = Sample Absorbance < Low Calibration Point		
Date	v RESULT v	std	DUPLICATE
12-16-96	Cr +6 µg/g =		
1350			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

Data Entry by:	<i>[Signature]</i>	Date:	12/16/96
Approved by:	<i>[Signature]</i>	Date:	12/16/96
Form 265101S2 Rev. 1.2		Page 1 of 1	

Cr+6 : LA-265-101 (B-0)

SOLID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.645	0.647
sample	Sample Volume in mL (Vsamp)	1.0 mL	1.0 mL
Work List	Method Factor (10 + Vsamp) (MF)		
0605	Dilution Factor (DF)		
Test Code	Digest g/L (DGL)	4.7030	4.6768
	Low calibration abs point	0.021	
Matrix	High calibration abs point	0.779	
SOLID	Y-intercept (a)	0.00337	
Sample #	Slope (b)	0.78285	
51610542.64	Cr +6 µg/g =		
Calibration Date			
12/13/96			
Instrument Code	Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)		
	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Analyst	Less Than Value = Sample Absorbance < Low Calibration Point		
10/20/96/10001	v RESULT v	sample	DUPLICATE
Date	Cr +6 µg/g =		
12-16-96			
Time			
13:55			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

Data Entry by:	<i>[Signature]</i>	Date:	12/16/96
Approved by:	<i>[Signature]</i>	Date:	11/1/97
Form 265101S2 Rev. 1.2		Page	1 of 1

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE		SPIKE
SPRKE 11119-1	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.046
	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.390
	SAMPLE VOLUME (Sample) in mL (Vsam)	1.0 mL
	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	
	VOLUME of SPIKE in mL (VOL Spk)	1.0 mL
	SPIKE DILUTION FACTOR (SDF)	
	SPIKE VALUE (µg/mL)	
	LOW CALIBRATION POINT	0.021
	HIGH CALIBRATION POINT	0.779
	Y-INTERCEPT (a)	0.00537
	SLOPE (b)	0.78285
Instrument Code	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSV)	
	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)	
	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)	
	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)	
	% Spike Rec. = (MSSV * TVSS) / SPIKE VALUE * VOL Spk / SDF	

RESULT % SPIKE RECOVERY = ERR

Date entry by: <i>[Signature]</i>	Date: 12/16/96
Approved by: <i>[Signature]</i>	Date: <i>[Signature]</i>

Cr+6 : LA-265-101 (B-0)

SOLID SAMPLE

Cr+6 : LA-265-101 (B-0) SOLID SAMPLE		sample	DUPLICATE
Type	Sample Absorbance at 540 nm- (abs)	0.046	0.044
sample 271054363	Sample Volume in mL (Vsamp)	1.0ml	1.0ml
Work List	Method Factor (10 + Vsamp) (MF)		
12305	Dilution Factor (DF)		
Test Code	Digest g/L (DGL)	5.3160	5.3060
	Low calibration abs point	0.021	
Matrix	High calibration abs point	0.779	
SOLID	Y-intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
271054363	Cr +6 µg/g =		

Calibration Date

12/13/96

Instrument Code

Cr +6 µg/g = (abs-a) * MF * DF * (1000mL/L) / (b * DGL * Vsamp)

RPD = |Cr +6 µg/g SAM - Cr +6 µg/g DUP| / AVERAGE OF THE SAM + DUP * 100

Less Than Value = Sample Absorbance < Low Calibration Point

Analyst

Lab Analyst

RESULT

Date

12-16-96

Time

13:50

	sample	DUPLICATE
Cr +6 µg/g		

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

Data Entry by	Date: 12/16/96
Approved by	Date: 12/16/96
Form 26510152 Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE			SPIKE
Type	SAMPLE ABSORBANCE at 540 nm	(SAM abs)	0.046
SPIKE IN 19-F	SAMPLE + SPIKE ABSORBANCE at 540 nm	(S+S abs)	0.290
Work List	SAMPLE VOLUME (Sample) in mL	(Vsam)	1.0 mL
1/23/96	SAMPLE VOLUME (Spiked Sample) in mL	(S+S Vsam)	
Test Code	VOLUME of SPIKE in mL	(VOL Spk)	1.0 mL
	SPIKE DILUTION FACTOR	(SDF)	
Matrix	SPIKE VALUE	(µg/mL)	
Sample #	LOW CALIBRATION POINT		0.021
541304363	HIGH CALIBRATION POINT		0.779
Calibration Date	Y-INTERCEPT	(a)	0.00537
12/13/96	SLOPE	(b)	0.78285
Instrument Code	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSV)		
Analyst	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)		
Volume (Mm) 11/16/96	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)		
Date	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)		
12-16-96	% Spike Rec. = (MSSV * TVSS) / SPIKE VALUE * VOL Spk / SDF		
Time			
12:50			

RESULT	% SPIKE RECOVERY	=	ERR

Date entry by: <i>[Signature]</i>	Date: 12/16/96
Approved by: <i>[Signature]</i>	Date: 12/16/96
Form 265101X3 Rev. 1.2	Page 1 of 1

LABCORE Data Entry Template for Worklist# 14499

Analyst: WLM Instrument: SPOT ^{BA001} _{WIA 10 23 96} Book # 11A19E

Method: LA-265-101 Rev/Mod B-0

Worklist Comment: CRVI-01 FOR AN-105 FB V#10755

RTS!

GROUP	PROJECT	S TYPE	SAMPLER	N A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			CRVI-01	LIQUID	<u>5.00</u>	<u>5.18</u>	<u>N/A</u>	ug/ml
		2 BLNK			CRVI-01	LIQUID	<u>1.0</u>	<u>5.220</u>	<u>N/A</u>	ug/ml
96000945	AN-105	3 SAMPLE	996T004003	0	CRVI-01	LIQUID	<u>N/A</u>	<u>5.419</u>	<u>419</u>	ug/ml
96000945	AN-105	4 DUP	996T004003	0	CRVI-01	LIQUID	<u>5.419</u>	<u>5.419</u>	<u>N/A</u>	ug/ml
96000945	AN-105	5 SPK	996T004003	0	CRVI-01	LIQUID	<u>1.00</u>	<u>1.074</u>	<u>N/A</u>	ug/ml

Final page for worklist # 14499

WLM 12-16-96
Analyst Signature Date

WLM 12-23-96
Analyst Signature Date

Date Entry & Validation

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

471

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		blank	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	
blank	Sample Volume in ml. (Vsamp)	1.00	
Work List	Method Factor (10 + Vsamp) (MF)	11	10
12803	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
S95T004303	Cr +6 µg/mL =	<0.220	ERR

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/95	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vim	v RESULT v	blank	DUPLICATE
Date	Cr +6 µg/mL =	<0.220	ERR
12/16/96			
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Date Entry by: <i>David Taylor</i>	Date: 12/19/96
Approved by: <i>David Taylor</i>	Date: 12/19/96

Form 265101LT Rev. 1

Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELLER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		std	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.374	
std	Sample Volume in mL (Vsamp)	1.00	
Work List	Method Factor (10 * Vsamp) (MF)	11	10
12803	Dilution Factor (DF)	1	
Test Code	Digest Dilution Factor (DDF)	1	
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00537	0.00537
Sample #	Slope (b)	0.78285	0.78285
S96T004003	Cr +6 µg/mL =	5.18E+00	ERR

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/96	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
sp01			
Analyst			
vln	v RESULT v	std	DUPLICATE
Date	Cr +6 µg/mL =	5.18E+00	ERR
12/16/96			
Time			

The Duplicate Absorbance is < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Data Entry by: <i>Jeanette Lopez</i>	Date: 12/19/96
Approved by: <i>Jeanette Lopez</i>	Date: 12/19/96
Form 285101LT Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.000	0.000
sample	Sample Volume in mL (Vsamp)	0.50	0.50
Work List	Method Factor (10 * Vsamp) (MF)	10.5	10.5
12803	Dilution Factor (DF)	1	1
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point	0.021	0.021
Matrix	High Calibration Point	0.779	0.779
LIQUID	Y-Intercept (a)	0.00637	0.00637
Sample #	Slope (b)	0.78285	0.78285
9561004003	Cr +6 µg/mL =	<0.419	<0.419

Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/98			
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP]*100		
sp01	Less Than Value = abs < Low Calibration Point		
Analyst			
vin	v RESULT v	sample	DUPLICATE
Date			
12/16/98	Cr +6 µg/mL =	<0.419	<0.419
Time			

Both Absorbances are < Detection Limit (the Low Calibration Point)

RESULT > Recovery % Difference = N/A

Data Entry by:	Date: 12/19/98
Approved by:	Date: 12/18/98
Form 265101LT Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0
SPIKE	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.371
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.500
14499	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.500
Test Code	VOLUME of SPIKE in mL (VOL Spk)	1.000
CRVI-01	SPIKE DILUTION FACTOR (SDF)	1
Matrix	SPIKE VALUE (µg/mL)	5
liquid		
Sample #	LOW CALIBRATION POINT	0.821
596T004003	HIGH CALIBRATION POINT	0.779
Calibration Date	Y-INTERCEPT (a)	0.00537
12/13/96	SLOPE (b)	0.78286
Instrument Code		
sp01	Cr+6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSGV)	
Analyst	Cr+6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)	
vim	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)	
Date	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)	
12/16/96	% Spike Rec. = (MSSV * TVSS) / SPIKE VALUE * VOL Spk / SDF	
Time		

RESULT	% SPIKE RECOVERY	=	107.4%
--------	------------------	---	--------

Data entry by: <i>[Signature]</i>	Date: 12/19/96
Approved by: <i>[Signature]</i>	Date: 12/19/96
Form 265101X3 Rev. 1.2	Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

	blank	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	
blank	Sample Volume in mL (Vsamp)	
Work List	Method Factor (10 + Vsamp) (MF)	
14499	Dilution Factor (DF)	
Test Code	Digest Dilution Factor (DDF)	
	Low Calibration Point	0.021
Matrix	High Calibration Point	0.779
LIQUID	Y-intercept (a)	0.00537
Sample #	Slope (b)	0.78285
5112204003	Cr +6 µg/mL =	

Calibration Date

12/13/96

Instrument Code

Analyst

Valerie M. Mason

Date

12-16-96

Time

$$\text{Cr +6 } \mu\text{g/mL} = (\text{abs-a}) * \text{MF} * \text{DF} * \text{DDF} / (\text{b} * \text{Vsamp})$$

$$\text{RPD} = ((\text{Cr +6 } \mu\text{g/mL SAM} - \text{Cr +6 } \mu\text{g/mL DUP}) / \text{AVERAGE OF THE SAM + DUP}) * 100$$

Less Than Value = abs < Low Calibration Point

v RESULT v

$$\text{Cr +6 } \mu\text{g/mL} =$$

blank

DUPLICATE

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT >

N/A

Data Entry by:

Date: 12/16/96

Approved by:

Date: 12/17

Form 265101LT Rev. 1.0

Page 1 of 1

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		std	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.3585	
std	Sample Volume in mL (V _{sample})	1.000	
Work List	Method Factor (10 * V _{sample}) (MF)		
Test Code	Dilution Factor (DF)		
Test Code	Digest Dilution Factor (DDF)		
	Low Calibration Point	0.021	
Matrix	High Calibration Point	0.778	
LIQUID	Y-Intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
1019-F	Cr +6 µg/mL =		
Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * V _{sample})		
12/13/96	RPD = ((Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP)*100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
Analyst	v RESULT v		
Date	std	DUPLICATE	
12-16-96	Cr +6 µg/mL =		
Time			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT > N/A

Date Entry by:	<i>[Signature]</i>	Date:	12/16/96
Approved by:	<i>[Signature]</i>	Date:	12/16/96

PLACE ANALYTICAL CARD HERE OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0)

LIQUID SAMPLE

		sample	DUPLICATE
Type	Sample Absorbance at 840 nm (abs)	0.000	0.000
sample	Sample Volume in mL (Vsamp)	20.0 mL	20.0 mL
Work List	Method Factor (10 * Vsamp) (MF)		
	Dilution Factor (DF)		
Test Code	Digest Dilution Factor (DDF)		
	Low Calibration Point	0.021	
Matrix	High Calibration Point	0.779	
LIQUID	Y-Intercept (a)	0.00537	
Sample #	Slope (b)	0.78285	
09/07/00 400-3	Cr +6 µg/mL =		
Calibration Date	Cr +6 µg/mL = (abs-a) * MF * DF * DDF / (b * Vsamp)		
12/13/95	RPD = ((Cr +6 µg/mL SAM - Cr +6 µg/mL DUP) / AVERAGE OF THE SAM + DUP) * 100		
Instrument Code	Less Than Value = abs < Low Calibration Point		
Analyst	v RESULT v	sample	DUPLICATE
12/13/95	Cr +6 µg/mL =		
Date			
12/13/95			
Time			

The Sample Absorbance is < Detection Limit (the Low Calibration Point)

RESULT > N/A

Data Entry By:	Date: 12/18/96
Approved by:	Date: 1/16/97

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

Cr+6 : LA-265-101 (B-0) HEXAVALENT CHROMIUM - SPIKE		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.000
SPIKE	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.511
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	10.0
1-1-11	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	10.0
Test Code	VOLUME of SPIKE in mL (VOL Spk)	1.000
	SPIKE DILUTION FACTOR (SDF)	
Matrix	SPIKE VALUE (µg/mL)	
Sample #	LOW CALIBRATION POINT	0.021
C-9100400-5	HIGH CALIBRATION POINT	0.779
Calibration Date	Y-INTERCEPT (a)	0.00537
12/13/86	SLOPE (b)	0.78265
Instrument Code	Cr +6 Molarity in SAMP VIAL = (SAM abs - a) / b (MSV)	
Analyst	Cr +6 Molarity in S+S VIAL = (S+S abs - a) / b (MSSV)	
12-16-86	TOTAL VOLUME for SAMP VIAL = 10 + Vsam (TVS)	
12-16-86	TOTAL VOLUME for S+S VIAL = 10 + S+S Vsam + VOL Spk (TVSS)	
Time	% Spike Rec. = (MSSV * TVSS) / SPIKE VALUE * VOL Spk / SDF	

RESULT	% SPIKE RECOVERY	ERR
--------	------------------	-----

Date entry by:	<i>[Signature]</i>	Date:	12/16/86
Approved by:	<i>[Signature]</i>	Date:	12/16/86

Form 265101X Rev. 1.2/1

Page 1 of 1

LABCORE Completed Worklist Report for Worklist# 10544

Analyst: cal

Instrument: IC02

Book# 149240

Method: 1A-533-10.5 Rev/Mod D-1

Worklist Comment: AN-105 IC. RCJ

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	QCB	0	0	010-QC	Y	QC	1	<1.36e-2	ug/mL
1	QCB	0	0	010-QC	CU	QC	1	2.12e-01	0.051 ug/mL
1	QCB	0	0	010-QC	MO2	QC	1	<1.07e-1	ug/mL
1	QCB	0	0	010-QC	BR	QC	1	<1.72e-1	ug/mL
1	QCB	0	0	010-QC	MO3	QC	1	1.66e-01	0.168 ug/mL
1	QCB	0	0	010-QC	PO4	QC	1	<1.15e-1	ug/mL
1	QCB	0	0	010-QC	SO4	QC	1	<1.35e-1	ug/mL
1	QCB	0	0	010-QC	CHLORATE	QC	1	<1.02e-1	ug/mL
1	QCB	0	0	010-QC	Y	QC	5.50e01	6.06e01	102.712 % Recovery
1	QCB	0	0	010-QC	CL	QC	7.59e01	6.18e01	105.038 % Recovery
1	QCB	0	0	010-QC	NO2	QC	8.42e02	5.72e02	105.720 % Recovery
1	QCB	0	0	010-QC	BR	QC	5.09e01	5.01e01	102.017 % Recovery
1	QCB	0	0	010-QC	MO2	QC	5.94e02	6.09e02	101.929 % Recovery
1	QCB	0	0	010-QC	PO4	QC	5.85e02	5.46e02	99.017 % Recovery
1	QCB	0	0	010-QC	SO4	QC	6.31e02	6.55e02	103.887 % Recovery
1	QCB	0	0	010-QC	CHLORATE	QC	5.25e02	5.57e02	108.005 % Recovery
2	SAMPLE	2547003631	0	010-QC	Y-03	L1Q010	N/A	1.44e+01	14.440 ug/mL
2	SAMPLE	2547003631	0	010-QC	CL-02	L1Q010	N/A	1.08e+01	10.090 ug/mL
2	SAMPLE	2547003631	0	010-QC	NO2-02	L1Q010	N/A	1.18e+02	119.900 ug/mL
2	SAMPLE	2547003631	0	010-QC	BR-01	L1Q010	N/A	2.427e+04	140.000 ug/mL
2	SAMPLE	2547003631	0	010-QC	PO3-02	L1Q010	N/A	2.392e+02	195.500 ug/mL
2	SAMPLE	2547003631	0	010-QC	PO4-01	L1Q010	N/A	1.321e+02	132.100 ug/mL
2	SAMPLE	2547003631	0	010-QC	SO4-01	L1Q010	N/A	1.530e+02	151.000 ug/mL
2	SAMPLE	2547003631	0	010-QC	CHLORATE	L1Q010	N/A	1.167e+02	116.700 ug/mL
4	DUP	2547003631	0	010-QC	Y-02	L1Q010	<1.44e0	<1.44e0	RPD
4	DUP	2547003631	0	010-QC	CL-02	L1Q010	<1.08e0	<1.08e0	RPD
4	DUP	2547003631	0	010-QC	NO2-01	L1Q010	<1.18e0	<1.18e0	RPD
4	DUP	2547003631	0	010-QC	BR-02	L1Q010	2.43e+04	2.41e+04	6.626 RPD
4	DUP	2547003631	0	010-QC	MO2-01	L1Q010	2.38e+02	2.09e+02	5.567 RPD
4	DUP	2547003631	0	010-QC	PO4-02	L1Q010	<1.32e0	<1.32e0	RPD
4	DUP	2547003631	0	010-QC	SO4-02	L1Q010	<1.53e0	<1.53e0	RPD
4	DUP	2547003631	0	010-QC	CHLORATE	L1Q010	<1.17e0	<1.17e0	RPD

Final page for worklist# 10544

Analyst Signature

Date

Analyst Signature

Date

James M. Joyce 8/28/96

Units shown for QC (BLN/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 10544

Analyst: ELM Instrument: IC00 2002 Book# 149N9-C

Method: LA-533-105 Rev/Mod D-1

Worklist Comment: AN-105 IC. RCJ

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 CCV			@IC-QC	QC		
3 SAMPLE	596T003631 0		@IC-01	LIQUID	96055945	AN-105
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
4 DUP	596T003631 0		@IC-01	LIQUID		

Final page for worklist # 10544

A. Lambert 0825-96
Analyst Signature Date

Analyst Signature Date

10544 AG.CSV

Chris Kell for GME

uploaded 8/23/96 GME

Data Entry Comments:

A-0010-IC				DATA FILE/WORKLIST RESOLUTION				28-Aug-96	
Worklist#: 10544				Data File: 10544AG.CSV					
	Seq	Type	Sample #	Seq#	Data File	Sample Name	Dilution		
-	=>	1 CCB		-	1 96082501.d17	INST. BLANK	1.00		
	=>	2 CCV			2 96082501.d18	149N9-C STD.	101.00		
	=>	3 SAMPLE	S96T003631		96082501.d19	S96T003631	10201.00		
	=>	4 DUP	S96T003631		3 96082501.d20	S96T003631	1111.00		
					4 96082501.d21	S96T003631 DUP	1111.00		
					96082501.d22	S96T003631 SPIKE	1111.00		
+				+					

Save(F4) Abort(Shift-F3) ListFiles(Shift-F1) UploadFile(F8)

Data Reprocessed On 08/26/1996 14:53:07

Sample Name: INST. BLANK Date: 08/25/1996 03:19:55
 Data File : C:\DX\DATA\96082501.d17
 Method : C:\DX\METHOD\ANIONS.MET
 ACl Address: 1 System: 1 Inject#: 17 Detector: CDM-1
 Analyst : *[Signature]* Column:

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 1 3000 5Hz 0.00 10.00 200

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.36		0.000	55	357	2	
4	1.65	CHLORIDE	0.031	76	352	2	0.81
5	3.58	NITRATE	0.168	37	200	1	2.78
6	6.68	SULFATE	0.116	36	355	1	2.99
Totals			0.315	203	1264		

File: 96082501.d17 Sample: INST. BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 484 TO 487

Data Reprocessed On 08/26/1996 14:53:09

Sample Name: 149N9-C STD Date: 08/25/1996 03:31:35
 Data File : C:\DX\DATA\96082501.d18
 Method : C:\DX\METHOD\ANTONS.MRT
 ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
 Analyst : Column:

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	1000	5Kz	0.00	10.00		200

***** Peak Report: All Peaks *****

PK. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	111	532	2	
2	1.05	FLUORIDE	60.610	2322	11719	2	0.96
3	1.65	CHLORIDE	81.363	2154	10683	1	0.81
4	2.00	NITRITE	572.902	9230	53711	1	0.00
5	3.06	BROMIDE	600.893	5429	36637	2	0.11
6	3.50	NITRATE	608.549	5083	48144	2	0.57
7	5.15	PHOSPHATE	544.127	1806	21146	1	2.39
8	6.85	SULFATE	652.829	4681	66525	1	2.49
9	9.07	OXALATE	557.217	2144	40960	1	3.34
Totals			3678.490	33561	290055		

File: 96082501.d18 Sample: 149N9-C STD.

Data Reprocessed On 12/06/1996 10:06:09

Sample Name: INST. BLANK Date: 11/24/1996 07:43:22
Data File : C:\DX\DATA\96112401.d01
Method : C:\DX\METHOD\ACE.MET
ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-1
Analyst *John F. [Signature]* Column: AG4A-SC, AS4A-SC, SRS

Calibration Volume *2.0* Dilution *1/6/97* Points Rate Start Stop Area Reject
96112401

External 1 1 4500 5Hz 0.00 15.00 0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. %Delta Code
1	5.73		0.000	2504	97890	2
2	7.23		0.000	2037	57063	2
3	8.16		0.000	267	1592	2
4	8.40		0.000	664	46962	2
Totals			0.000	5472	203507	

File: 96112401.d01 Sample: INST. BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES *986* TO *995*

DISTRIBUTION SHEET

To Distribution	From Production Planning & Control	Page 1 of 1
		Date: 01/23/97
Project Title/Work Order HNF-SD-WM-DP-199, Rev. 1, "Tank 241-AN-105, Cores 152 and 153 Analytical Results for the Final Report"		EDT NO.: N/A
		ECN NO.: ECN-633339

Name	MSIN	Text With all Attach	EDT/ECN ONLY	
------	------	-------------------------	-----------------	--

DE&S Hanford, Inc.

J. E. Meacham	S7-14	X		
---------------	-------	---	--	--

U.S. Department of Energy, RL

C. A. Babel	S7-54	X		
-------------	-------	---	--	--

Lockheed Martin Hanford Corp.

J. N. Appel	G3-21		X	
J. Jo	R2-12	X		
T. J. Kelley	S7-21		X	
N. W. Kirch	R2-11	X		
M. J. Kupfer	H5-49		X	
J. B. Schaffer	R2-12		X*	
L. A. Stauffer	R2-11	X		
J. A. Voogd	H5-03		X	
A. E. Young	R2-12	X		

Lockheed Martin Services, Inc.

Central Files	A3-88	2		
---------------	-------	---	--	--

Los Alamos Technical Associates

M. T. Ellsworth	T6-06		X*	
-----------------	-------	--	----	--

Pacific Northwest National Laboratory

J. R. Gormsen	K7-28		X	
S. J. Harris	K7-22	X		
K. L. Silvers	K9-08		X	

Rust Federal Services of Hanford, Inc.

M. W. Barnes	T6-16	X		
D. C. Hetzer	H5-27		X	
K. L. Powell	T6-04		X	
C. M. Seidel	T6-04	X		
F. H. Steen	T6-06	X		
LTIC	T6-03		X	

SGN Eurisys Services Corp.

B. A. Higley	H5-27		X	
--------------	-------	--	---	--

Defense Nuclear Federal Safety Board

Rich Tontodonato		X		
------------------	--	---	--	--

625 Indiana Ave. N. W.
Washington D.C. 20004

* Needs only releasing paperwork, not a copy of the released document.

HNF-SD-WM-DP-199, REV. 1

Worklist # 14431 pu
Analyst MV (Printed initials)

Analysis Date 11-6-96
Analysis Time 1500

Matrix (SOLID) (LIQUID) (EACH) ← Circle one

Tracer Book # 126343[illegible]

HNF-SD-WM-DP-199, REV. 1

8

AT : LA-508-101 (D-2) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	8	9
11612	COUNT TIME IN MINUTES (CT)	30	30
Att or TB ?	BACKGROUND in cpm (BKG)	0.17	0.17
AT	SAMPLE SIZE in mL (SS)	0.250	0.250
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2104	0.2104
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.296	0.337
Batch Number			
96007946	Blank Concentration in $\mu\text{Ci/L}$	< 2.58E+01	
Regrun	Replicate Concentration in $\mu\text{Ci/L}$	< 2.94E+01	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 2.9398E+01	
Sample Prep			
N/A	$R_s (\text{Sample Count Rate}) = (TC / CT) - BKG$		
Sample #	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S96T003975	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SLF	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 2.94E-02	DETECTION LEVEL
Analyst			
TAM	LESS THAN Value was Determined from Rmax.		
Date Complete			3.86E-02
08/12/96	RELATIVE COUNTING ERROR	244.6%	$\mu\text{Ci/mL}$
Analysis Date			
08/08/96			
Analysis Time			
03:00 AM			
Sample Point			
AN-105			

Analyst:	TAM	Date: 12-Aug-96
Signature of Chemist:	SLF	Date: 1/14/97
SAMPLE.WB1 Rev. 1.0	508101ML	

Sample Name: 144N9-B STD. Date: 11/24/1996 08:03:01
 Data File : C:\DX\DATA\96112401.d02
 Method : C:\DX\METHOD\ACE.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector:CDM-1
 Analyst : Column: AG4A-SC,AS4A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4500	5Hz	0.00	15.00		0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	2.10	fluoride	66.909	1956	13931	2	0.00
2	2.37	acetate	122.877	707	5002	2	0.99
3	3.11	formate	121.776	1361	9537	1	2.19
4	5.87		0.000	4308	243930	2	
5	7.23		0.000	1751	68108	2	
6	8.16		0.000	273	1619	2	
7	8.41		0.000	678	46693	2	
Totals			311.562	11035	388819		

File: 96112401.d02 Sample: 144N9-B STD.

Data Reprocessed On 08/30/1996 14:17:18

```

=====
| Sample Name: S96T003751 SPIKE                      Date: 08/19/1996 00:04:06 |
| Data File  : E:\DATA\96081861.D07                  |
| Method     : C:\DX\METHOD\ANIONS.MET              |
| ACI Address: 1 System: 1 Inject#: 7                 |
| Analyst    :                               Column:   |
|                                                    |
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          4141   3000  5Hz   0.00 10.00      200
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.88		0.000	661	3410	1	
2	1.06	FLUORIDE	988.406	988	4609	1	1.60
3	1.54	CHLORIDE	12436.302	6910	39869	2	-1.28
4	1.87	NITRITE	139482.456	53748	332860	2	-1.06
5	2.76	BROMIDE	23922.446	6014	35413	2	0.98
6	3.02	NITRATE	180858.007	43137	413448	2	0.11
7	4.55	PHOSPHATE	23750.655	1326	16799	1	0.89
8	5.89	SULFATE	25309.131	4866	62949	1	0.91
9	7.70	OXALATE	21156.856	2208	38067	1	1.58
Totals			427904.259	119859	947423		

File: 96081861.D07 Sample: S96T003751 SPIKE

Sample Name: S96T004448 DUP Date: 10/08/1996 11:10:11
 Data File : C:\DX\DATA\96100801.D05 WHC-SD-WM-DP-199, REV. 1
 Method : C:\DX\METHOD\ACE.MET
 ACI Address: 1 System: 2 Inject#: 5 Detector: CDM-1
 Analyst : Ed Colvin Column: AG4A-SC, AS4A-SC, SRS

HNF-SD-WM-DP-199, REV. 1

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 11 4320 5Hz 0.00 14.40 0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	2.13	fluoride	7.368	2524	13191	2	0.16
2	2.40	acetate	4.233	229	1338	2	-1.23
3	2.55		0.000	200	1199	2	
4	3.11	formate	5.931	547	3911	1	-0.75
5	3.95		0.000	275	2183	1	
6	5.40		0.000	13881	110784	2	
7	5.76		0.000	42593	661785	2	
8	7.41		0.000	36446	721753	3	
9	8.15		0.000	1260	9073	4	
10	8.35		0.000	2203	175220	4	
Totals			17.532	100159	1700435		

File: 96100801.D05 Sample: S96T004448 DUP

08/28/96 14:11

A-0004-1

LABCORE Data Entry Template for Worklist# 12463Analyst: DCD Instrument: AB00 IS Book# 47B56Method: LA-508-101 Rev/Mod D-2

Worklist Comment: AN-105.Determine sample size using Iudium.Spike:0.1ml. new

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1 STD				@ALPHA01	SOLID		
2 BLNK-PREP				@ALPHA01	SOLID		
3 BLNK/BKG				@ALPHA01	SOLID		
4 SAMPLE	S96T004317	0	F	@ALPHA01	SOLID	96000945	AN-105
	Analytes Requested: ALPHA01 , ALPHA01E						
5 DUP	S96T004317	0	F	@ALPHA01	SOLID		
6 SPK	S96T004317	0	F	@ALPHA01	SOLID		
7 SAMPLE	S96T004337	0	F	@ALPHA01	SOLID	96000945	AN-105
	Analytes Requested: ALPHA01 , ALPHA01E						
8 DUP	S96T004337	0	F	@ALPHA01	SOLID		
9 SPK	S96T004337	0	F	@ALPHA01	SOLID		

Final page for worklist # 12463Analyst Signature David Dunham Date 8-29-96Analyst Signature C. J. O'Brien Date 8/30/96
Sharon L. Lulick 8-30-96

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1586

INTERIM

Table 1. Data Summary Results
AN-105

CORE NUMBER: 153
SEGMENT #: 8

SEGMENT PORTION: Drainable Liquid

Sample #	R #	Analyte	Unit	Action Limits		Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S961003975		DSC Exotherm using Mettler	Joules/g	Lower	Upper	96.31	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003975		DSC Exotherm DRY Calculated	Joules/g Dry	-1.0e+03	480.0	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S961003975		% Water by TGA using Mettler	%	None	None	99.48	n/a	49.99	50.16	50.08	0.34	n/a	n/a	n/a
S961003975		Alpha in Liquid Samples	UCI/mL	-1.0e+03	6.5E08	93.75	<2.94e-02	<2.94e-02	<1.45E-1	n/a	n/a	90.25	3.90e-02	2.45E+02

⇒ Limit Violated
⇒ Selected Limit

WHC-SD-WM-DP-199, REV. 0

HNF-SD-WM-DP-199, REV. 1

WHC-SD-WM-DP-199, REV. 1

2-77