RMIS View/Print Document Cover Sheet

This document was retrieved from the Boeing ISEARCH System.

Accession #: D196054850

Document #: SD-WM-RPT-146

Title/Desc:

VAPOR & GAS SAMPLING OF SST 241TY101 USING THE
VAPOR SAMPLING SYSTEM

SEP 20 1995

ENGINEERING DATA TRANSMITTAL

Page 1 of 1
1. EDT No 612713

2. To: (Receiving Organization) DISTRIBUTION		3. From: (Originating Organization) SAMPLING AND MOBILE LABS OM624		4. Related EDT No.: N/A	
5. Proj./Prog./Dept./Div.: VAPOR/HTS/SAMPLING AND MOBILE LABS		6. Cog. Engr.: RICKY MAHON 3-7437		7. Purchase Order No.: N/A	
8. Originator Remarks: 241-TY-101, SAMPLING USING THE VAPOR SAMPLING SYSTEM				9. Equip./Component No.: N/A	
				10. System/Bldg./Facility: M0292/WSCF	
11. Receiver Remarks:				12. Major Assm. Dwg. No.: N/A	
				13. Permit/Permit Application No.: N/A	
				14. Required Response Date: 17 JULY 95	

15. DATA TRANSMITTED					(F)	(G)	(H)	(I)
(A) Item No.	(B) Document/Drawing No.	(C) Sheet No.	(D) Rev. No.	(E) Title or Description of Data Transmitted	Approval Designator	Reason for Transmittal	Originator Disposition	Receiver Disposition
1	WCH-SD-WM-RPT-146 WHC- KMB 9/2/95	ALL	0	VAPOR AND GAS SAMPLING OF SINGLE-SHELL TANK 241-TY-101 USING THE VAPOR SAMPLING SYSTEM	Q	2	1	

16. KEY					
Approval Designator (F)		Reason for Transmittal (G)		Disposition (H) & (I)	
E, S, Q, D or N/A (see WHC-CM-3-5, Sec.12.7)		1. Approval 2. Release 3. Information	4. Review 5. Post-Review 6. Dist. (Receipt Acknow. Required)	1. Approved 2. Approved w/comment 3. Disapproved w/comment	4. Reviewed no/comment 5. Reviewed w/comment 6. Receipt acknowledged

(G)	(H)	17. SIGNATURE/DISTRIBUTION (See Approval Designator for required signatures)								(G)	(H)
Reason	Disp.	(J) Name	(K) Signature	(L) Date	(M) MSIN	(J) Name	(K) Signature	(L) Date	(M) MSIN	Reason	Disp.
2	1	Cog. Eng. R.D. MAHON	<i>R D Mahon</i>	20/4/95	S3-27	RESOURCE CENTER			R2-12	3	
2	1	Cog. Mgr. J.D. DORIAN	<i>J D Dorian</i>	9/5/95	S3-27						
2	1	QA C.J. STEPHAN	<i>C J Stephan</i>	9/4/95	H4-16						
3		DAVE BRATZEL			S7-21						
3		JOHN DEICHMAN			T6-03						
3		RAMPUR VISWANATH			S3-90						
3		G.S.T.T.	<i>KMB</i>	9/2/95	A3-36						

18. GLENN S. CAPRIO <i>Glenn S. Caprio</i> 7/28/95 Signature of EDT Originator Date		19. JOHN J. DORIAN <i>John J. Dorian</i> 9/5/95 Authorized Representative for Receiving Organization Date		20. COGNIZANT MANAGER <i>John J. Dorian</i> 9/5/95 Cognizant Manager Date		21. DOE APPROVAL (if required) Ctrl. No. ☐ Approved ☐ Approved w/comments ☐ Disapproved w/comments	
---	--	---	--	---	--	---	--

THIS PAGE INTENTIONALLY
LEFT BLANK

RELEASE AUTHORIZATION

Document Number: WHC-SD-WM-RPT-146, REV 0

Document Title: Vapor and Gas Sampling of Single-Shell Tank 241-TY-101 Using the Vapor Sampling System

Release Date: 9/20/95

**This document was reviewed following the
procedures described in WHC-CM-3-4 and is:**

APPROVED FOR PUBLIC RELEASE

WHC Information Release Administration Specialist:

Kara Broz

TRADEMARK DISCLAIMER. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

This report has been reproduced from the best available copy. Available in paper copy. Printed in the United States of America. To obtain copies of this report, contact:

Westinghouse Hanford Company - Document Control Services
P.O. Box 1970, Mailstop H6-08, Richland, WA 99352
Telephone: (509) 372-2420; Fax: (509) 376-4989

THIS PAGE INTENTIONALLY
LEFT BLANK

SUPPORTING DOCUMENT

1. Total Pages 35 36 KMB
9/25/95

2. Title

VAPOR AND GAS SAMPLING OF SINGLE-SHELL TANK 241-TY-101 USING THE VAPOR SAMPLING SYSTEM

3. Number

WHC-SD-WM-RPT-146

4. Rev No.

0

5. Key Words

241-TY-101
VSS
SUMMA
TST
SORBENT
TANK

6. Author

Name: G.S. CAPRIO

Glenn S Caprio
Signature

Organization/Charge Code 0M624 / E61920

7. Abstract

THIS DOCUMENT PRESENTS SAMPLING DATA RESULTING FROM THE APRIL 6, 1995, SAMPLING OF SST 241-TY-101 USING THE VAPOR SAMPLING SYSTEM.

8. RELEASE STAMP

OFFICIAL RELEASE
BY WHC
DATE SEP 20 1995

Sta. 4

THIS PAGE INTENTIONALLY
LEFT BLANK

APPROVALS

Document title: Vapor and Gas Sampling of Single-Shell Tank 241-TY-101
Using the Vapor Sampling System

Approved by:

R. A. Westberg, Field Scientist
Vapor Sampling Project
Sampling and Mobile Laboratories

6/20/95
Date

Approved by:

R. D. Mahon, Project Lead
Vapor Sampling Project
Sampling and Mobile Laboratories

20 AUG 95
Date

Approved by:

D. L. Edwards, Technical Lead
Vapor Sampling Project
Sampling and Mobile Laboratories

6/22/95
Date

Approved by:

R. S. Viswanath, Principal Scientist
Vapor Sampling Project
Special Analytical Studies

8/29/95
Date

THIS PAGE INTENTIONALLY
LEFT BLANK

CONTENTS

1.0	SCOPE	1
2.0	SAMPLING EQUIPMENT DESCRIPTION	1
2.1	VAPOR SAMPLING SYSTEM	1
3.0	SAMPLING EVENT DESCRIPTION	3
3.1	SPECIFICATIONS	3
3.2	OPERATIONS AND SAMPLING PERSONNEL	3
3.3	INDUSTRIAL HYGIENE FIELD RESULTS	3
3.4	AMBIENT CONDITIONS	4
3.5	SAMPLE COLLECTION	4
3.6	FIELD GC/FID RESULTS	5
3.7	RADIATION SCREENING	5
4.0	SAMPLE CHAIN OF CUSTODY: RECEIPT, STORAGE, AND SHIPMENT	7
5.0	QUALITY ASSURANCE AND CONTROLS	9
5.1	VAPOR SAMPLING SYSTEM CLEANING	9
5.2	INSTRUMENT CALIBRATION	9
6.0	ANOMALIES	10
7.0	REFERENCES	11
APPENDICES		
A	SAMPLE LOG SHEETS	A-1
B	AMBIENT CONDITIONS	B-1
C	CHAIN-OF-CUSTODY FORMS	C-1

LIST OF TABLES

1	Flow Control Calibration	2
2	TOC Results	5
3	Radionuclide Analysis Results	6
4	Oak Ridge National Laboratory Samples	7
5	Pacific Northwest Laboratory Samples	8
6	Calibration Data	10

LIST OF TERMS

CGI	Combustible Gas Indicator
COC	Chain Of Custody
DOT	U.S. Department of Transportation
GC/FID	Gas Chromatograph/Flame Ionization Detector
GEA	Gamma Energy Analysis
HEPA	High-Efficiency Particulate Air (filter)
NH ₃	Ammonia
NO ₂	Nitrogen Dioxide
NO	Nitric Oxide
H ₂ O	Water Vapor
OPC	Offsite Property Control
ORNL	Oak Ridge National Laboratory
OVM	Organic Vapor Meter
PNL	Pacific Northwest Laboratory
SML	Sampling and Mobile Laboratories
SST	Single-Shell Tank
TCP	Tank Characterization Plan
team	SML Vapor Team
TST	Triple Sorbent Trap
VSS	Vapor Sampling System
WHC	Westinghouse Hanford Company

This page intentionally left blank.

VAPOR AND GAS SAMPLING OF SINGLE-SHELL TANK 241-TY-101 USING THE VAPOR SAMPLING SYSTEM

1.0 SCOPE

The Vapor Issue Resolution Program tasked the Sampling and Mobile Laboratories (SML) to collect representative headspace samples from Hanford Site single-shell tank (SST) 241-TY-101. This document presents sampling data resulting from the April 6, 1995 sampling of SST 241-TY-101. Analytical results will be presented in separate reports issued by the laboratories that supplied and analyzed the sampling media.

2.0 SAMPLING EQUIPMENT DESCRIPTION

2.1 VAPOR SAMPLING SYSTEM

The SML vapor team (the team) used the vapor sampling system (VSS) to collect representative samples of the air, gases, and vapors from the headspace of SST 241-TY-101 on April 6, 1995. Mahon et al. (1994) describes in detail the VSS, its performance, and its operation. The team used the VSS to collect sorbent trap and SUMMA¹ canister headspace samples from SST 241-TY-101. The team then sent these samples to the analytical laboratories for analysis.

The VSS comprises a mobile laboratory, a hot-water-jacketed stainless steel probe that is inserted into the tank headspace, and stainless steel transfer tubing that connects the mobile laboratory to the probe. A vacuum pump draws sample vapor from the tank headspace and through all transfer tubing and the sampling manifold. Electrically activated, pneumatically actuated, valves direct sample flow within the VSS. Instrumentation housed in the mobile laboratory monitors and controls system temperature, monitors absolute and differential system pressure, meters and controls vapor mass flow, and monitors sample vapor total organic carbon (TOC) content using a gas chromatograph/flame ionization detector (GC/FID).

A key feature of the VSS is its use of heated transfer tubing and a heated sampling manifold. Maintaining the system temperature at an electronically controlled elevated temperature prevents vapor condensation and reduces vapor adsorption on surfaces exposed to sample vapor. Mahon et al. (1994) describes various tests and observations that indicate the VSS sample transfer efficiency is consistently high.

Sorbent trap samples are collected at the sorbent trap station of the sampling manifold. Sorbent traps are pencil-size stainless steel or glass tubes that

¹SUMMA is a registered trademark of Molectrics, Inc., Cleveland, Ohio.

contain vapor-adsorbing media. A known amount of sample vapor is passed through the tube, which traps (by adsorption) virtually all the target analytes. The concentration of analytes in the vapor sampled is calculated from the quantity of analyte found in the sorbent media and the volume of vapor passed through the sorbent trap.

The sorbent trap station uses highly accurate mass flow controllers to measure and control the flow rate of sample vapor through the sorbent traps. The controllers are located downstream of the sorbent trap station and the in-line driers, which remove moisture from the sample vapor before it is metered. Errors associated with the mass flow controllers were determined by the Westinghouse Hanford Company (WHC) Standards Laboratory before the SST 241-TY-101 sampling event (see Table 1). Flow rates and the duration of flow are specified by the analytical laboratories that supply and analyze the sorbent traps.

Table 1. Flow Control Calibration.

Flow-indicating Control Valve	Flow (stdcm ³ /min)	Error (%)
1	200	±0.7
3	200	±0.05

The VSS is also equipped with a gas chromatograph (GC). The Hewlett Packard 5890 Series II GC is equipped with a flame-ionization detector (FID), a 1 mL sample loop, a 10 port injection valve, a 2 meter chromatographic column, a programmable oven, and a portable computer loaded with the HP-Chemstation software used to control the GC. The GC is plumbed to directly transfer sample from the VSS manifold to the GC sample loop. After the sample is transferred into the sample loop and reaches equilibrium, the run is initiated manually. HP-Chemstation software activates the 10 port valve to transfer the sample from the sample loop to the column. The sample passes through the column and the FID generates a signal for total organic carbon. All data is then transmitted to the computer where it is stored for further analysis.

The GC is equipped with a HP-5 column which is 2 meters long, 0.25 mm inside diameter, and which contains a 0.25 um phenyl methyl silicone phase. The GC oven is programmed to heat from 50 °C to 270 °C at a rate of approximately 70 °C per minute. Helium is the carrier gas, air and hydrogen the combustion gases, and nitrogen the make-up gas.

The GC/FID is configured to quantitatively estimate concentrations of total organic carbon. The GC/FID confirms sampling system cleanliness, ambient air background TOC concentrations during sampling, and TOC concentration of tank vapor samples. The system is multi-point calibrated at the weather station on a as available basis, the last time being January 1995. The GC/FID has displayed a high degree of stability over a period of months. For further details, see Section 3.6, Field GC/FID Results.

The sampling manifold also has a station for sampling vapor with evacuated SUMMA™ canisters. SUMMA™ canisters are stainless steel vessels with internal surfaces that have been prepared by the SUMMA™ process, which passivates active sites on the canister walls to minimize adsorption of gases and vapors. An analytical laboratory must clean and evacuate SUMMA™ canisters before use. The evacuated canister is filled with sample vapor through a manually operated valve, which is then closed to seal the sample inside. SUMMA™ canisters essentially allow collection and transfer of whole-air samples from the sample site to an analytical laboratory where the sample is analyzed. The standard capacity of each SUMMA™ canister is 6 liters.

3.0 SAMPLING EVENT DESCRIPTION

3.1 SPECIFICATIONS

The Vapor Issue Resolution Program specifies sampling requirements in WHC-SD-WM-TP-299, *Tank 241-TY-101 Tank Characterization Plan* (Homi 1995). The Tank Characterization Plan (TCP) also specifies the types and number of samples to be collected, flow rates, and durations. These key sampling parameters are summarized on the sample log sheets in Appendix A. In addition to the sample log sheets, checksheets for each individual sample help ensure correct sampling procedures. SML retains these documents in the project file. This sample event's project-specific number is S5-019.

3.2 OPERATIONS AND SAMPLING PERSONNEL

Glenda Biever was the Tank Farm Operations person-in-charge. The other SML vapor team members included:

- R. A. Westberg, Field Scientist
- R. L. Ragan, Field Scientist
- N. P. Buechler, VSS Technician
- T. B. Utecht, VSS Technician.

The VSS was set up at SST 241-TY-101 on April 5, 1995 and was allowed to warm up overnight. Sampling began shortly after 9:00 a.m. on April 6, 1995, and was completed by 2:30 p.m. the same day.

3.3 INDUSTRIAL HYGIENE FIELD RESULTS

Prior to hooking up to SST 241-TY-101, an industrial hygiene technician field tested tank vapors. The technician purged the vapor probe sample tube for 5 minutes and then field measured vapor stream contents using a combustible gas indicator (CGI) and an organic vapor meter (OVM). The measurements were verbally reported, LEL 0.0% and the total organic carbon content of 0.8 ppm. The technician also measured 10 ppm for NH₃.

3.4 AMBIENT CONDITIONS

The weather the day of the sampling event, April 6, 1995 was sunny and clear with little winds. Graphs of ambient temperatures and pressures taken at the Hanford Meteorological Station, which is about 2 miles west of TY Farm, are provided in Appendix B.

3.5 SAMPLE COLLECTION

The hot-water-jacketed sampling probe was located in Riser 3 of SST 241-TY-101. The probe length, from the sample inlet to the top of the riser flange, was 5.5 meters.

All zones of the VSS were heated to 40 °C during setup of the VSS at SST 241-TY-101 on April 5, 1995. The team stabilized the VSS temperature zones by 9:00 a.m. on April 6, 1995, and the system was ready to collect samples. Measured according to the VSS operating procedure, the pressure and temperature of SST 241-TY-101 were 983 mbar (737.3 torr) and 15.6 °C, respectively. The sample log sheets (Appendix A) provide a complete chronology of the sample event including start and end times, flow rates, volumes, and specific sample identifiers.

Approximately 21.5 hours before the first samples were collected, the team began heating the VSS transfer tubing and sampling manifold. During this warmup period, the team began a purge of all vapor transfer tubing and the sampling manifold with ambient air. Prior to sampling tank vapors on April 6, 1995, the team collected two SUMMA™ canister samples of ambient air, one manually 10 meters upwind of the VSS connection with SST 241-TY-101, and the other using the VSS sampling manifold. The former was collected to establish background levels of trace organic vapors, and the latter was collected to establish the cleanliness of the sampling manifold.

A leak check of the VSS sampling manifold and transfer tubing (up to the connection to the sampling probe) was performed prior to opening the tank probe vapor sampling valve. The system was evacuated to 287 mbar (215.4 torr) and leakage of ambient air into the system was observed by monitoring system pressure for 15 minutes. Leakage resulted in an increase of 4.3 mbar (3.2 torr) in system pressure during the 15 minute test. Given a system volume of not more than 10 L, this pressure increase corresponds to a leak rate of approximately 9.9 mL/min at a 287 mbar (215.4 torr). This leak rate was then estimated for average SUMMA™, TST, and sorbent sampling pressures. It was found that for the SUMMA™ canisters, dilution by ambient air was approximately 0.1%, for TST traps sampled at 50 mL/min the dilution was approximately 0.3%, for TST traps sampled at 200 mL/min the dilution was approximately 0.4%, and for sorbent traps the dilution was approximately 0.3%.

The tank probe vapor sampling valve was opened and the VSS was purged with sample vapor from SST 241-TY-101 for 30 minutes at a total flow rate of 5.46 L/min. This purge was performed to flush ambient air from the system and saturate the system's active adsorption sites. Because the volume of transfer

tubing and the sampling manifold upstream of the sampling devices is estimated to be no more than 10 L, this purge provided about 16.4 air turnovers in the system.

Two analytical laboratories provided sample media. Oak Ridge National Laboratory (ORNL) provided triple sorbent traps (TSTs) for organic vapors; and Pacific Northwest Laboratory (PNL) provided SUMMA™ canisters and sorbent traps for organic vapors, ammonia (NH₃), nitrogen dioxide (NO₂), nitric oxide (NO), and water vapor (H₂O).

3.6 FIELD GC/FID RESULTS

The GC was single point calibrated on the day of sampling using 6.0 ppmC Propane standard by Scott Specialty Gas. The standard is an E.P.A. Protocol Gas that is $\pm 2\%$ NIST traceable.

Table 2 displays the Field GC/FID results from the sampling of 241-TY-101.

Table 2. TOC Results.

Number of Runs	Description	Average TOC Concentration (ppmC)	% Standard Deviation
4	Propane	6.00	1.19
4	Ambient	2.69	0.66
6	Tank Vapor	3.43	1.58

3.7 RADIATION SCREENING

Samples are unconditionally released from the SST farm in accordance with 1995-33300-RSP-03, *Release of Vapor Sampling Equipment* (WHC 1995c). Radiological screening results are used to determine (1) if the samples must be shipped as radioactive or nonradioactive in accordance with U.S. Department of Transportation (DOT) regulations and (2) if the samples meet the laboratory acceptance criteria.

The DOT limits for shipping a nonradioactive sample are 2000 combined pCi/g of beta-gamma activity and alpha activity. Samples exceeding these DOT limits may be shipped as radioactive material if the samples do not exceed the following laboratory acceptance criteria:

ORNL: Beta-gamma activity <450 pCi/g of sample media. Alpha activity <135 pCi/g.

PNL: Beta-gamma activity <400 pCi/g of sample media. Alpha activity <100 pCi/g of sample media.

To protect the sampling manifold and sampling devices from radioactive particulates, all sample vapor for the April 6, 1995, SST 241-TY-101 vapor sampling event was drawn through a series of four glass-fiber high-efficiency particulate air (HEPA) filters placed upstream of the sampling manifold. These four filters were in place any time tank vapors were flowing through the system. When sampling was complete, the filters were removed and assigned unique sample identifiers. All four HEPA filters were submitted to Laboratory 222-S for total alpha, total beta, and gamma energy analysis (GEA). The HEPA filter closest to the sampling manifold was analyzed to determine if the samples met DOT shipping criteria and laboratory acceptance criteria. Moisture from the tank vapors was collected in a silica gel trap through one of the sorbent station ports and analyzed for tritiated water. SST 241-TY-101 filter and silica gel analysis results are shown in Table 3. SML scientists use the activity results in Table 3 to calculate pCi/g of sample media. SML maintains this information in the project specific file. The results in Table 3 indicate that the samples collected from SST 241-TY-101 met the laboratory acceptance criteria and the DOT definition of a nonradioactive shipment.

Table 3. Radionuclide Analysis Results.

Filter	Sample Identifier	Activity Results ^a (pCi/sample)	Activity ^b (pCi/L of tank gas)
Upstream HEPA filter (box)	S5019-A35.0U1	Total Alpha = 4.91 Total Beta = 13.4 GEA = <detectable	= 0.015 = 0.04 = <detectable
Downstream HEPA filter (box)	S5019-A36.0D1	Total Alpha = <1.42 Total Beta = 2.84 GEA = <detectable	= <0.004 = 0.009 = <detectable
Upstream HEPA filter (VSS)	S5019-A37.0U2	Total Alpha = 1.62 Total Beta = <1.60 GEA = <detectable	= 0.005 = <0.005 = <detectable
Downstream HEPA filter (VSS)	S5019-A38.0D2	Total Alpha = 1.93 Total Beta = 17.8 GEA = <detectable	= 0.006 = 0.05 = <detectable
Tritium trap	S5019-A03.0T1	Total activity = <50.0	= <50.0 ^c

NOTES:

The samples are nonradioactive. These results were evaluated against Laboratory acceptance criteria and DOT limits.

^aAll less than (<) values represent the minimum detection limits at Laboratory 222-S.

^bNumbers based on an approximation of the total volume of tank vapor through the HEPA filters. Appendix A and the sample checksheets were used to estimate a total flow through the VSS of 334 L.

^cNumber is calculated using a total volume of 1 L passing through the tritium trap.

4.0 SAMPLE CHAIN OF CUSTODY: RECEIPT, STORAGE, AND SHIPMENT

All sorbent trains, sorbent tubes, and SUMMA™ canisters received by SML are kept in a custody locked storage area maintained by SML. Sorbent trains and tubes were maintained at 4 ± 2 °C in a refrigeration unit. SUMMA™ canisters were stored in the same locked storage area, but were not refrigerated.

TSTs are supplied by ORNL and ORNL initiates the chain-of-custody (COC) forms. The sample media is placed in 40-mL volatile organic analysis vials. Evidence tape is applied to the vials, which are then placed in a shipping container along with the relinquished COC forms and shipped to SML. TSTs are shipped to SML in a cooler containing "blue" ice.

The TSTs were received, inventoried against the COC forms, and the "received by" block on the COC form was signed by SML. They were maintained at 4 ± 2 °C before and after the sampling event in the storage area refrigeration unit.

The TSTs were shipped back to ORNL after the sampling event using offsite property control (OPC W95-0-0304-34). Table 4 lists the sample identifiers, sample types, and COC form numbers for all ORNL samples. The samples were delivered to shipping on April 17, 1995 with instructions to ship via Federal Express, next-day delivery.

Table 4. Oak Ridge National Laboratory Samples.

Sample Identifier	Sample Type	COC Number
S5019-A7.912	TST	006762
S5019-A8.913	TST	006762
S5019-A9.914	TST	006762
S5019-A10.910	TST field blank	006762
S5019-A11.915	TST	006762
S5019-A12.916	TST	006762
S5019-A13.917	TST	006762
S5019-A14.918	TST	006762
S5019-A15.919	TST	006762
S5019-A16.911	TST	006762
S5019-A17.926	TST field blank	006762
S5019-A18.920	TST	006762
S5019-A19.921	TST	006762
S5019-A20.922	TST	006762
S5019-A21.924	TST trip blank	006762
S5019-A22.925	TST trip blank	006762

PNL supplied sorbent chain traps and SUMMA™ canisters, including COC forms. These Sampling devices were picked up from PNL by SML and transported in a government vehicle to a custody locked storage area. The sorbent trains were maintained at 4 ± 2 °C before and after the sampling event in storage area refrigeration unit.

After sampling, the PNL sorbent traps and SUMMA™ canisters were transported by government vehicle directly to PNL and delivered to J. A. Edwards on April 13, 1995. Table 5 lists the sample identifiers, sample types, and COC form numbers for all PNL samples.

Table 5. Pacific Northwest Laboratory Samples.

Sample Identifier	Sample Type	COC Number
S5019-A01.024	Ambient upwind SUMMA™	008884
S5019-A02.027	Ambient SUMMA™ (VSS)	008884
S5019-A04.072	SUMMA™	008884
S5019-A05.075	SUMMA™	008884
S5019-A06.076	SUMMA™	008884
S5019-A23.54U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A24.55U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A25.56U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A26.57U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A27.58U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A28.59U	NH ₃ /NO _x /NO/H ₂ O sorbent	008885
S5019-A29.60U	NH ₃ /NO _x /NO/H ₂ O trip blank	008885
S5019-A30.61U	NH ₃ /NO _x /NO/H ₂ O trip blank	008885
S5019-A31.62U	NH ₃ /NO _x /NO/H ₂ O trip blank	008885
S5019-A32.63U	NH ₃ /NO _x /NO/H ₂ O field blank	008885
S5019-A33.64U	NH ₃ /NO _x /NO/H ₂ O field blank	008885
S5019-A34.65U	NH ₃ /NO _x /NO/H ₂ O field blank	008885

From the time that samples are received by SML until they are shipped back to the analytical laboratory, all COCs are maintained by SML in accordance with WHC-IP-1127-1.3, *Chain-of-Custody/Special Analysis Request for RCRA and CERCLA Protocol Samples*, (WHC 1995a). Copies of the completed COC forms for this sampling event are included in Appendix C.

5.0 QUALITY ASSURANCE AND CONTROLS

5.1 VAPOR SAMPLING SYSTEM CLEANING

Immediately prior to sampling of SST 241-TY-101, the VSS manifold and vapor sample transfer tubing was heated to 60 °C and ambient air was purged through the system for 22.1 hours. Also, all pertinent system valves were actuated to release any contaminants that may have collected in the VSS valves themselves. After this purge an ambient air sample was drawn through the VSS manifold and a GC/FID run was initiated. No significant level of contaminants was detected. A second ambient air GC/FID run confirmed that the VSS manifold and transfer tubing was free of organic remnant residue down to ambient levels. A SUMMA[™] canister ambient air sample was then collected to confirm by laboratory analysis that the VSS sampling manifold was free of trace organic contaminants (or to determine which contaminants were present and at what concentration), as discussed in Section 3.5. For further details, refer to Appendix C of WHC-IP-1127-4.5 (WHC 1995b) and the project-specific file located with SML.

The hot-water-jacketed probe was initially cleaned in compliance with LO-080-405/A0, *Vapor Space Probe Cleaning Procedure* (WHC 1993) before it was installed in SST 241-BY-112. This procedure requires a solvent rinse of all internal probe surfaces with acetone and methanol to clear the sample line of possible contamination remaining from the construction process. The probes are then heated to 90 °C, and dry air is passed through the probe to evaporate the solvents. Before the probes are released, the team ensures via a hand-held organic vapor meter that the total organic concentration in the sample line is below 1 ppmv. Following the sampling of 241-BY-112 the probe was removed and transferred to TY Farm. Before installation in 241-TY-101 the probe was recleaned in accordance with Appendix C of WHC-IP-1127-4.5 (WHC 1995b).

5.2 INSTRUMENT CALIBRATION

Instruments located in the VSS are calibrated on an annual basis at the WHC Standards Laboratory. VSS instrumentation calibration data, maintained in files by SML, are summarized in Table 6. According to the calibration schedule shown in Table 6, all instrumentation was within its calibration period during the SST 241-TY-101 sampling event.

Table 6. Calibration Data.

System	Calibration Date	Expiration Date	WHC Standards Laboratory Code
Temperature Control System	03/30/95	03/30/96	804-67-74-009
Mass Flow System	12/08/94	12/08/95	804-28-03-001
Pressure System	12/07/94	12/07/95	804-67-89-001

5.3 BLANK SAMPLES

Trip blanks are samples that accompany the sample media from the point of generation through sample analysis. They are transported to the field with the sample collection media but remain unopened during the sampling event. Analysis of trip blanks is used to assess cross-contamination of sample media during field transport and storage.

Field blanks are sampling devices similar to trip blanks. They are prepared and handled in the same manner as the sampling media, including mounting on the sampling port for 1 minute, but no tank vapors are drawn through the field blanks.

Spiked blanks are prepared as regular sampling media but also contain a known amount of special analyte. Tank vapors are drawn through these blanks and they are handled and analyzed just like any other sample. Analysis of the spiked blanks is used to evaluate potential sample loss during shipment or storage.

Ambient blanks are samples of ambient air collected at the sampling location. Analysis of ambient blanks is used to assess contamination that may be present in the atmosphere or in the transfer tubing or sampling manifold of the VSS immediately prior to sampling operations.

Tables 4 and 5 also list sample blanks used during the sampling of SST 241-TY-101.

6.0 ANOMALIES

All samples were collected in accordance with the TCP and WHC-IP-1127-4.5, *Collection of SUMMA Canisters and Sorbent Tube Samples Using the Vapor Sampling System* (WHC 1995b).

7.0 REFERENCES

- 49 CFR 100-177, 1992, "Transportation," *Code of Federal Regulations*, as amended.
- Mahon, R. D., C. M. Jones, and M. S. Story, 1994 (draft), *Evaluation of the Capabilities and Use of the Vapor Sampling System for Tank Headspace Sampling and Characterization*, SD-WM-RPT-094, Westinghouse Hanford Company, Richland, Washington.
- Homi, C. S., 1995, *Tank 241-TY-101 Tank Characterization Plan*, WHC-SD-WM-TP-299, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1995a, *Chain-of-Custody/Special Analysis Request for RCRA and CERCLA Protocol Samples*, Procedure WHC-IP-1127-1.3, Rev. 1, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1995b, *Collection of SUMMA Canisters and Sorbent Tube Samples Using the Vapor Sampling System (VSS)*, Procedure WHC-IP-1127-4.5, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1995c, *Release of Vapor Sampling Equipment*, Procedure 1995-33300-RSP-03, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1993, *Vapor Space Probe Cleaning Procedure*, Procedure LO-080-405/A0, Westinghouse Hanford Company, Richland, Washington.

This page intentionally left blank.

APPENDIX A
SAMPLE LOG SHEETS

This page intentionally left blank.

Waste Tank
241-TY-101

Set up VSS (Section A)(Temperature set point = 40°C)

Ensure HEPA filters are installed

Ensure connection to sample probe

System status check sheet. (Verify zones are to temp)

Tank Temperature 15.6 °C

GC Standard Runs

GC Ambient Air Runs through port 10

PNL

OGI/ORNL Port

Sample ID Valve #

WHC Sample ID

Group # 1

Description

Flow

Rate

Desired

Duration

(min.)

Desired

Flow

Rate

Desired

Flow

Rate

Desired

Flow

Rate

Desired

Flow

Rate

Desired

Flow

Rate

Desired

Flow

Rate

Desired

Flow

Date: April 6, 1995

VSS Personnel:

R. Westberg, R. Ragan

T. Utecht

N. Beuchler

Trailer Personnel:

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

N. Beuchler

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Time

4/6/95

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Total

Flow

Liters

6947.66

Actual

Start

Time

4/5/95

4/6/95

Actual

End

Waste Tank
241-TY-101

WHC Sample ID	PNL OGI/ORNL Sample ID	Port Valve #	Description	Desired Flow Rate	Desired Duration (min.)	Desired Total Flow	Actual Flow Rate	Actual Start Time	Actual End Time	Actual Total Time	Actual Total Flow
S5019- A07. 912	<u>TST#912</u>	9	TST #1	<u>50</u>	<u>4</u>	200	<u>50.40</u>	<u>1058</u>	<u>1102</u>	<u>4</u>	<u>0.20</u>
S5019- A08. 913	<u>TST#913</u>	10	TST #2	<u>50</u>	<u>4</u>	200	<u>50.40</u>	<u>1058</u>	<u>1102</u>	<u>4</u>	<u>0.20</u>
S5019- A09. 914	<u>TST#914</u>	8	TST #3	<u>50</u>	<u>4</u>	200	<u>50.50</u>	<u>1111</u>	<u>1115</u>	<u>4</u>	<u>0.20</u>
S5019- A11. 915	<u>TST#915</u>	10	TST #4	<u>50</u>	<u>4</u>	200	<u>50.50</u>	<u>1111</u>	<u>1115</u>	<u>4</u>	<u>0.20</u>
S5019- A10. 910	<u>TST#910</u>		TST FIELD BLANK #1					<u>1111</u>	<u>1112</u>	<u>1</u>	
S5019- A12. 916	<u>TST#916</u>	9	TST # 5	<u>200</u>	<u>5</u>	1000	<u>200.40</u>	<u>1123</u>	<u>1128</u>	<u>5</u>	<u>1.00</u>
S5019- A13. 917	<u>TST#917</u>	10	TST # 6	<u>200</u>	<u>5</u>	1000	<u>200.40</u>	<u>1123</u>	<u>1128</u>	<u>5</u>	<u>1.00</u>
S5019- A14. 918	<u>TST#918</u>	8	TST # 7	<u>200</u>	<u>5</u>	1000	<u>200.40</u>	<u>1135</u>	<u>1140</u>	<u>5</u>	<u>1.00</u>
S5019- A15. 919	<u>TST#919</u>	10	TST # 8	<u>200</u>	<u>5</u>	1000	<u>200.40</u>	<u>1135</u>	<u>1140</u>	<u>5</u>	<u>1.00</u>
S5019- A16. 911	<u>TST#911</u>	9	TST # 9	<u>200</u>	<u>20</u>	4000	<u>200.40</u>	<u>1149</u>	<u>1209</u>	<u>20</u>	<u>4.01</u>
S5019- A18 920	<u>TST#920</u>	10	TST # 10	<u>200</u>	<u>20</u>	4000	<u>200.40</u>	<u>1149</u>	<u>1209</u>	<u>20</u>	<u>4.01</u>
S5019- A17 926	<u>TST#926</u>		TST FIELD BLANK #2					<u>1150</u>	<u>1151</u>	<u>1</u>	
S5019- A19. 921	<u>TST#921</u>	8	TST # 11	<u>200</u>	<u>20</u>	4000	<u>200.40</u>	<u>1225</u>	<u>1245</u>	<u>20</u>	<u>4.01</u>
S5019- A20. 922	<u>TST#922</u>	10	TST # 12	<u>200</u>	<u>20</u>	4000	<u>200.40</u>	<u>1225</u>	<u>1245</u>	<u>20</u>	<u>4.01</u>
GC Run # 5 (Tank run #3)								<u>1253</u>			

Waste Tank
241-TY-101

WHC Sample ID	PNL OGI/ORNL	Port	Description	Desired Flow Rate	Desired Duration (min.)	Desired Total Flow	Actual Flow Rate	Actual Start Time	Actual End Time	Actual Total Time	Actual Total Flow
S5019- A23. 54U	54U	9	NH3/NOx/H2O #1	200	15	3000	200.40	1258	1313	15	3.01
S5019- A24. 55U	55U	10	NH3/NOx/H2O #2	200	15	3000	200.40	1258	1313	15	3.01
S5019- A32. 63U	63U	8	3/NOx/H2O FIELD BLANK #1					1258	1259	1	
S5019- A25. 56U	56U	8	NH3/NOx/H2O #3	200	15	3000	200.40	1320	1335	15	3.01
S5019- A26. 57U	57U	10	NH3/NOx/H2O #4	200	15	3000	200.40	1320	1335	15	3.01
S5019- A33. 64U	64U	9	3/NOx/H2O FIELD BLANK #2					1320	1321	1	
GC Run # 6 (Tank run # 4)								1340			
S5019- A27. 58U	58U	9	NH3/NOx/H2O #5	200	15	3000	200.40	1344	1359	15	3.01
S5019- A28. 59U	59U	10	NH3/NOx/H2O #6	200	15	3000	200.40	1344	1359	15	3.01
S5019- A34. 65U	65U	10	3/NOx/H2O FIELD BLANK #3					1404	1405	1	
GC Run # 7 (Tank run #5)								1404			
GC Run #8 (Tank run # 6)								1410			

TOTAL TANK GAS USED DURING SAMPLING RUNS

57.88

S5019- A35. 0U1	T-1883 (222S)	Upstream HEPA(box)
S5019- A36. 0D1	T-1884 (222S)	Downstream HEPA(box)
S5019- A37. 0U2	T-1885 (222S)	Upstream HEPA (VSS)
S5019- A38. 0D2	-1886 (222S)	Downstream HEPA (VSS)

1420

1420

1420

1420

				Waste Tank 241-TY-101							
WHC Sample ID	PNL	Port Valve #	Description	Desired	Desired	Desired	Actual	Actual	Actual	Actual	Actual
	OGI/ORNL			Flow	Duration	Total	Flow	Start	End	Total	Total
	Sample ID			Rate	(min.)	Flow	Rate	Time	Time	Time	Flow
Trip Blanks (DO NOT EXPOSE)											
S5019- A21. 924	TST#924		TST TRIP #1								
S5019- A22. 925	TST#925		TST TRIP #1								
S5019- A29. 60U	<u>60U</u>		NH3/NOx/H2O TRIP #1								
S5019- A30. 61U	<u>61U</u>		NH3/NOx/H2O TRIP #2								
S5019- A31. 62U	62U		NH3/NOx/H2O TRIP #3								

A-6

WHC-SD-WM-RPT-146, Rev. 0

TY-101
TOTAL TANK VAPOR USED

	NUMBER	TIME	VOLUME	TOTAL		
	OF	IN	IN	VOLUME,		
	EVENTS	MINUTES	SCCM	LITERS		
LEAK CHECKS	1	0	5600	5.60		
TANK PURGE PUMP DOWNS	3	1	5600	16.80		
TANK PURGE TIME (From Spreadsheet)	1	30	5450	163.80		
GC PURGES	6	2	5000	60.00		
SUMMA PURGES	3	2	5000	30.00		
ALL SAMPLES COLLECTED				57.88		
TOTAL FOR TANK SAMPLING RUN				334.08		

A-7

WHC-SD-WM-RPT-146, Rev. 0

8-14

APPENDIX B
AMBIENT CONDITIONS

This page intentionally left blank.

WEATHER DURING VSS SAMPLING OF 241-TY-101

This page intentionally left blank.

APPENDIX C
CHAIN-OF-CUSTODY FORMS

This page intentionally left blank.

ORNL	CHAIN OF CUSTODY (Rev 1)		# 006762
Custody Form Initiator	Amy Dindal	Sampling Location	241-TY-101 (VSS)
ORNL Contact	Roger Jenkins	Telephone	(615) 576-8594
Method of Shipment	Federal Express	Shipped to	Rick Mahon
Sample Identification			

WHC Identifier	TST #	Trap Type	Received at WHC (✓)	Received at ORNL (✓)
S5019-A7.912	912	SAMPLE	✓	✓
S5019-A8.913	913	SAMPLE	✓	✓
S5019-A9.914	914	SAMPLE	✓	✓
S5019-A11.915	915	SAMPLE	✓	✓
S5019-A12.916	916	SAMPLE	✓	✓
S5019-A13.917	917	SAMPLE	✓	✓
S5019-A14.918	918	SAMPLE	✓	✓
S5019-A15.919	919	SAMPLE	✓	✓
S5019-A16.911	911	SAMPLE	✓	✓
S5019-A18.920	920	SAMPLE	✓	✓
S5019-A19.921	921	SAMPLE	✓	✓
S5019-A20.922	922	SAMPLE	✓	✓
S5019-A21.924	924	TRIP BLANK	✓	✓
S5019-A22.925	925	TRIP BLANK	✓	✓
S5019-A10.910	910	FIELD BLANK	✓	✓
S5019-A17.926	926	FIELD BLANK	✓	✓

COMMENTS: Collection Date 04-06-95
 Field Log Book # WHC-N-467-95
 TNA

Received @ ORNL OK

[] Field Transfer of Custody			Chain of Possession		(Sign and Print Names)	
Relinquished By	Date	Time	Received By	Date	Time	
AB Dindal / AB Dindal	3/30/95	747	T-BURET / T-BURET	03-31-95	1430	
T-BURET / T-BURET	04-06-95	0730	NT Burett / NT Burett	4-06-95	0730	
T-BURET / T-BURET	4-17-95	1000	AB Dindal / AB Dindal	4-18-95	1448	

Cooler Temperature Status

Temperature (°C)	Min	Max
When Shipped from ORNL	-3°	5°
When Received at WHC	-4°	8°
When Shipped from WHC	-4°	9°
When Received at ORNL	-2°	7°

Westinghouse
Hanford Company

CHAIN OF CUSTODY

WHC 008884

Custody Form Initiator J. A. Edwards - PNL

Telephone (509) 373-0141

Page 85-3009 / FAX 376-0418

Company Contact R. D. Mahon - WHC

Telephone (509) 373-2891

Page 85-3152 / FAX 373-3793

Project Designation/Sampling Locations 200 West Tank Farm
241-TY-101 Tank Vapor Sample SAF S5-019
(VSS Truck)

Collection date 04 - 06 - 95

Preparation date 03 - 30 - 95

Ice Chest No.

Field Logbook No. WHC- ⁶⁹⁷ ~~N-447~~ 10

Bill of Lading/Airbill No. N/A

Offsite Property No. N/A

Method of Shipment Government Truck

Shipped to PNL

Possible Sample Hazards/Remarks Unknown at time of sampling

Sample Identification

S5-019 - A01 .024

Ambient Air SUMMA #1 Upwind of TY-101

S5-019 - A02 .027

Ambient Air SUMMA #2 Through Port 15

S5-019 - A04 .072

SUMMA #3 Port 11

S5-019 - A05 .075

SUMMA #4 Port 13

S5-019 - A06 .076

SUMMA #5 Port 15

Field Transfer of Custody			Chain of Possession (Sign and Print Names)		
Relinquished By	Date	Time	Received By	Date	Time
J A Edwards <i>J A Edwards</i>	03-31-95	1340	T B Utech <i>T B Utech</i>	03-31-95	1340
<i>Tim Buehler / T B Utech</i>	04-06-95	0730	<i>NP Buehler / T B Utech</i>	4-6-95	0730
<i>NP Buehler / T B Utech</i>	4-13-95	1245	<i>J A Edwards / T B Utech</i>	4-13-95	1245

Final Sample Disposition

Comments:

PNL (only) Checklist

- 0 Media labeled and checked?
- 0 Letter of instruction?
- 0 Media in good condition?
- 0 COC info/signatures complete?
- 0 Rad release stickers on samples?
- 0 Activity report from 222S?
- 0 COC copy for LRB, RIDS filed?
- 0 COC copy for sorbent follow-on?

POC

Pick-up / Delivery

- (Y) N
- (Y) N
- (Y) N / (Y) N
- (Y) N / (Y) N
- (Y) N / (Y) N
- (Y) N / (Y) N
- (Y) N / (Y) N
- (Y) N / (Y) N

POC

C-4

Comments:

(Revised 10/17/94 PNL)

Westinghouse
Hanford Company

CHAIN OF CUSTODY

WHC 008885

Custody Form Initiator

J. A. Edwards - PNL

Telephone (509) 373-0141

Page 85-3009 / FAX 376-0418

Company Contact

R. D. Mahon - WHC

Telephone (509) 373-2891

Page 85-3152 / FAX 373-3793

Project Designation/Sampling Locations 200 West Tank Farm

241-TY-101 Tank

Vapor Sample SAF S5-019

(VSS Truck)

Collection date 04 - ~~06~~ - 95

Preparation date 03 -30 - 95

Ice Chest No.

Field Logbook No. WHC- ~~N~~ ⁶⁴⁷ - ~~547~~ ⁷⁸⁴ - 10

Bill of Lading/Airbill No.

N/A

Offsite Property No. N/A

Method of Shipment

Government Truck

Shipped to

PNL

Possible Sample Hazards/Remarks Unknown at time of sampling

Sample Identification

S5-019 - A23 . 54U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 1)	Line # 9
S5-019 - A24 . 55U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 2)	Line # 10
S5-019 - A25 . 56U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 3)	Line # 8
S5-019 - A26 . 57U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 4)	Line # 10
S5-019 - A27 . 58U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 5)	Line # 9
S5-019 - A28 . 59U -	NH ₃ /NO _x /H ₂ O (Sorbent Trap # 6)	Line # 10
S5-019 - A29 . 60U -	NH ₃ /NO _x /H ₂ O (Trap Trip Blank # 1)	
S5-019 - A30 . 61U -	NH ₃ /NO _x /H ₂ O (Trap Trip Blank # 2)	
S5-019 - A31 . 62U -	NH ₃ /NO _x /H ₂ O (Trap Trip Blank # 3)	
S5-019 - A32 . 63U -	NH ₃ /NO _x /H ₂ O (Field Blank # 1)	Line # 8
S5-019 - A33 . 64U -	NH ₃ /NO _x /H ₂ O (Field Blank # 2)	Line # 9
S5-019 - A34 . 65U -	NH ₃ /NO _x /H ₂ O (Field Blank # 3)	Line # 10

[] Field Transfer of Custody			[] Chain of Possession (Sign and Print Names)		
Relinquished By	Date	Time	Received By	Date	Time
G W Dennis <i>G W Dennis</i>	03-31-95	1030	J A Edwards <i>JA Edwards</i>	03-31-95	1030
J A Edwards <i>JA Edwards</i>	03-31-95	1340	T B Utecht <i>T B Utecht</i>	03-31-95	1340
T B Utecht <i>T B Utecht</i>	04-06-95	0120	R Brubaker <i>R Brubaker</i>	4-6-95	0230
R Brubaker <i>R Brubaker</i>	4-13-95	1245	JA Edwards <i>JA Edwards</i>	4-13-95	1245

Final Sample Disposition

Comments:

PNL (only) Checklist	Pick-up / Delivery	Comments:
Media labeled and checked?	☒ Y / ☐ N	
Letter of instruction?	☒ Y / ☐ N	
Media in good condition?	☒ Y / ☐ N	
COC info/signatures complete?	☒ Y / ☐ N	
Sorbents shipped on ice?	☒ Y / ☐ N	
Rad release stickers on samples?	☒ Y / ☐ N	
Activity report from 222S?	☐ Y / ☒ N	
COC copy for LRB, RIDS filed?	☐ Y / ☒ N	
COC copy for sorbent follow-on?	☐ Y / ☒ N/A	

POC

POC

(Revised 10/17/94 PNL)

Westinghouse Hanford
Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

COPY

009124

Page 1 of 1

Data Turnaround

☐ Priority
☒ Normal

Collector <i>T.B. ULTECHT</i>	Company Contact <i>R.D. Mahan</i>	Telephone No. <i>373-2891</i>
Project Designation <i>TY 101</i>	Sampling Location <i>241 TY 101</i>	SAF No. <i>SS019</i>
Ice Chest No. <i>N/A</i>	Field Logbook No. <i>WHC-N-467-10</i>	Method of Shipment <i>Via Gov't Vehicle #929</i>
Shipped To <i>220-S Labs 200 West</i>	Offsite Property No. <i>N/A</i>	Bill of Lading/Air Bill No. <i>N/A</i>
Possible Sample Hazards/Remarks <i>NONE</i>	Preservative <i>N/A</i>	
	Type of Container <i>Petri dish + Silica Gel</i>	
	No. of Container(s) <i>5</i>	
Special Handling and/or Storage <i>NONE</i>	Volume <i>47mL</i>	

SAMPLE ANALYSIS

Sample No.	Matrix*	Date Sampled	Time Sampled																	
SS019-0T1	Silica	4-6-95		T-1882																
SS019-0U1	Hepa Filter	4-6-95		T-1883																
SS019-0D1	Hepa Filter	4-6-95		T-1884																
SS019-0U2	Hepa Filter	4-6-95		T-1885																
SS019-0D2	Hepa Filter	4-6-95		T-1886																

CHAIN OF POSSESSION

Sign/Print Names

Relinquished By <i>AT Swahler</i>	Date/Time <i>4-7-95 0900</i>	Received By <i>[Signature]</i>	Date/Time <i>4-7-95 0900</i>
Relinquished By	Date/Time	Received By	Date/Time
Relinquished By	Date/Time	Received By	Date/Time
Relinquished By	Date/Time	Received By	Date/Time

SPECIAL INSTRUCTIONS

Please fax info to Tim Utech
*at 373-7076**Thanks**X = Silica Gel / Hepa Filters*

Matrix*

S	= Soil
SE	= Sediment
SO	= Solid
SL	= Sludge
W	= Water
O	= Oil
A	= Air
DS	= Drum Solids
DL	= Drum Liquids
T	= Tissue
WI	= Wipe
L	= Liquid
V	= Vegetation
X	= Other

LABORATORY
SECTION

Received By

Title

Date/Time

FINAL SAMPLE
DISPOSITION

Disposal Method

Disposed By

Date/Time

DISTRIBUTION: Original - Sample Yellow - Sampler