C/ORNL93-0218

ORNL/M--4998

**OAK RIDGE
NATIONAL
LABORATORY**

LOCKHEED MARTIN

Diffraction Null Compensators for Aspheric Testing

L. C. Maxey
J. E. Rogers
S. M. Arnold

Approved for public release;
distribution is unlimited.

MANAGED BY
LOCKHEED MARTIN ENERGY SYSTEMS, INC.
FOR THE UNITED STATES
DEPARTMENT OF ENERGY

UCN-13673 (38 6-95)

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

MASTER

This report has been reproduced directly from the best available copy.

Available to DOE and DOE contractors from the Office of Scientific and Technical Information, P.O. Box 62, Oak Ridge, TN 37831; prices available from (423) 576-8401, FTS 626-8401

Available to the public from the National Technical Information Service, U.S. Department of Commerce, 5285 Port Royal Rd., Springfield, VA 22161.

NTIS price codes-Printed Copy: A03 Microfiche A01

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Instrumentation and Controls Division

DIFFRACTIVE NULL COMPENSATORS FOR ASPHERIC TESTING

L. C. Maxey
J. E. Rogers
Oak Ridge National Laboratory

Steven M. Arnold
Diffraction International Ltd.

This report was developed for the Diffraction International Ltd.
Under DOE CRADA No. Ornl93-0218

Prepared by the
OAK RIDGE NATIONAL LABORATORY
Oak Ridge, Tennessee 37831-6285
managed by
LOCKHEED MARTIN ENERGY RESEARCH CORP.
for the
U.S. DEPARTMENT OF ENERGY
under contract DE-AC05-96OR22464

CONTENTS

ABSTRACT	v
1. OBJECTIVES OF THE CRADA	1
2. ASSURANCE THAT THE CRADA OBJECTIVES HAVE BEEN MET	1
3. BENEFITS TO DOE AND OAK RIDGE NATIONAL LABORATORY	2
4. TECHNICAL DISCUSSION OF WORK ACCOMPLISHED	2
5. INTELLECTUAL PROPERTY CONSIDERATIONS	3
6. ASSESSMENT OF COMMERCIALIZATION POSSIBILITIES	3
7. PLANS FOR FUTURE COLLABORATION	3
8. CONCLUSIONS	3
Appendix A: AFFORDABLE ASPHERIC TESTING TECHNOLOGY COMES TO LIGHT ..	5
Appendix B: COOPERATIVE RESEARCH BRINGS ADVANCED OPTICAL TESTING TECHNOLOGY TO MARKET AND TO DOE MISSIONS	11
Appendix C: FIGURE METROLOGY OF DEEP GENERAL ASPHERICS USING A CONVENTIONAL INTERFEROMETER WITH CGH NULL	15
Appendix D: DIFFRACTION INTERNATIONAL, LTD. PRODUCT LITERATURE AND ADVERTISEMENT COPY	29

ABSTRACT

This CRADA has combined the resources of a national laboratory and an innovative small company to evaluate and refine a revolutionary system for testing aspheric optics. This system uses special computer-generated holograms (CGHs), known as diffractive null correctors (DNCs), to achieve versatile and reliable measurements of aspheric optical components. The hardware is specifically designed to be compatible with existing test equipment common to most optical manufacturers.

Traditionally, the high costs associated with testing aspheric components have significantly limited their use. Through the activities of this CRADA Diffraction International has been able to commercialize a novel testing system that enables optics manufacturers to test aspheric optical components using their existing interferometric test equipment. As a result, the benefits of aspheric optical systems (high performance, compact size, and light weight) can be realized in new applications where cost has previously been prohibitive.

The Diffraction International "CGH Null Adapter" is available to the optics manufacturing industry at an attractive price. This product effectively reduces the entry level cost of testing aspheric optics with CGHs by an order of magnitude. The low cost hardware is backed with the support of world class expertise in the design and application of CGHs for aspheric testing. The combination will significantly advance the state of the art in aspheric optical testing equipment available to optics manufacturers.

1. OBJECTIVES OF THE CRADA

Aspheric optical surfaces have long been recognized for their ability to improve system performance and reduce system complexity. The overwhelming expense of making and testing them, however, has made these advantages irrelevant for most applications. In recent years, computer-controlled manufacturing techniques such as single point diamond turning and computer-controlled polishing have reduced aspheric manufacturing costs. Unfortunately, testing technology has lagged behind and remains a significant cost driver.

The development of computer generated holograms (CGHs) removed much of the uncertainty associated with testing aspheric optics. This technology is now considered the most reliable in the world for testing aspheric optics. CGH technology was, after all, relied upon for the final test of the corrective optics for the Hubble Space Telescope. A stronger testimony to the value of this technology cannot be conceived.

Despite their reliability, the use of CGHs for aspheric testing has been limited to a very few organizations with the necessary technical and financial resources to apply the technology. Previous implementations of the technology have required purchases of expensive specialized systems or expensive modifications of existing systems. The primary objective of this CRADA has been to make the technology accessible to and affordable for the optics manufacturing sector at large, by providing the technology in an inexpensive implementation compatible with existing test equipment.

2. ASSURANCE THAT THE CRADA OBJECTIVES HAVE BEEN MET

All of the tasks described under the Scope of Work in the original Statement of Work (SOW) have been performed. The original scope of work was exceeded, to the extent that three optics were tested (rather than two as stated in the SOW) and it became possible to repeat all of the CGH tests and corroborative tests using a second (different type) interferometer.

The CGH Null Adapter was evaluated and enhanced through this CRADA and is available as a commercialized product that is presently being used at the Oak Ridge National Laboratory. The ORNL personnel involved have previously evaluated every technology available for testing aspheric optics, as part of a large program to reduce the cost of manufacturing Strategic Defense Initiative (SDI) optics. During that multi-year SDI-sponsored ORNL program, reliable CGH testing technology could not be implemented for under \$100,000. By commercializing this once esoteric technology, the CGH Null Adapter reduces the entry level cost of reliable CGH testing to under \$10,000.

For a more detailed discussion of the significance of this CRADA and the objectives achieved, see Appendix A, "Affordable Aspheric Testing Comes to Light."

3. BENEFITS TO DOE AND OAK RIDGE NATIONAL LABORATORY

DOE has benefitted from this CRADA by participating in the commercialization of a technology that will significantly reduce the cost of aspheric optics manufacturing. The resulting benefits to DOE will accrue in programs requiring the use of aspheric optical elements. The demonstration of this technology within ORNL has enabled the technology to be tested and validated for application to existing and emerging DOE programs. These benefits are best explained in a previously generated document, Appendix B, "Cooperative Research Brings Advanced Optical Testing Technology to Market and to DOE Missions."

4. TECHNICAL DISCUSSION OF WORK ACCOMPLISHED

Closely following the task statements in the SOW, the CRADA was executed in a manner that culminated in a very complete conference publication (Appendix C, "Figure Metrology of Deep General Aspherics Using a Conventional Interferometer with CGH Null"). The type of optics to be tested and testing philosophy were jointly determined. The first two optics were diamond turned paraboloids, one a commercially available off-axis segment, the other an on-axis paraboloid manufactured by Oak Ridge National Laboratory. Mounting hardware for the optics was prepared and the CGHs for the two optics were designed.

The mounting and alignment hardware for the CGH components was designed and manufactured, the CGHs were mounted and aligned after they were manufactured. The optics were tested first using the CGH Null Adapter and then using autocollimating test configurations. The autocollimating test makes the paraboloid the easiest asphere to test without a null lens and is the reason that paraboloids were selected as the test optics for this CRADA. After completing the tests with the first interferometer, a Twyman Green configuration (manufactured by Breault Research Organization), they were tested with a second system, a Fizeau configuration (manufactured by Zygo).

Test results were analyzed jointly as the tests were conducted. In some cases discoveries were made during the testing that significantly affected the design or application of future CGH correctors. A novel technique for aligning off-axis aspheres to the CGH Null Adapter was developed at Oak Ridge (based partially on observations developed during alignment of previous off-axis systems). There was a discovery of adverse seasonal temperature effects (imposed by the Minnesota winter) on shipped optics - this resulted in a change in manufacturing processes that corrected the problem. Identification of the desired diffracted order and spatial filtering to eliminate the undesired diffracted orders were topics that resulted in greater product understanding, modified CGH corrector design philosophy, and an appreciation of performance variations between different optical test instruments.

During evaluation of the test results it was determined that testing a third optic, a custom made off-axis paraboloidal segment, would provide essential insight into the performance of the CGH Null Adapter with off-axis segments. This optic had been previously characterized and determined to have minimal errors, making comparison of data sets more meaningful. Because the tests of the first two optics had proceeded without any significant technical problems, it was determined that this optic could be added to the test set without additional cost.

The details of the CGH Null Adapter design and application, CGH corrector designs, comparison of test results and tolerance analysis were presented at the SPIE Annual Meeting and published in the proceedings (No. SPIE 2536-13). This detailed treatment of the CRADA

activities can be found in Appendix C, a pre-print of the SPIE conference proceedings paper "Figure Metrology of Deep General Aspherics Using a Conventional Interferometer with CGH Null."

5. INTELLECTUAL PROPERTY CONSIDERATIONS

There are patents pending on the CGH Null Adapter. These are the intellectual property of Diffraction International, Ltd., and were filed prior to the formation of the CRADA. No proprietary data were supplied to ORNL during the course of the CRADA. No protected CRADA information developed during the course of this CRADA, and no inventions were developed by either party as a result of the CRADA activities.

6. ASSESSMENT OF COMMERCIALIZATION POSSIBILITIES

The CGH Null Adapter is commercially available and is being supplied to industry. See Diffraction International, Ltd. product literature in Appendix D.

7. PLANS FOR FUTURE COLLABORATION

Multiple additional interactions between the CRADA partners have been discussed. This includes the possibility of an additional CRADA to develop and demonstrate methods of improving test accuracy by compensating or eliminating error sources.

An additional area of interest for future collaboration concerns the use of the CGH Null Adapter as a wavefront shaping component for the production of complex traditional holographic optics. This would make it possible to manufacture true holographic optical elements that could not be generated by computer aided manufacturing or other existing technology.

Cross-collaboration with other industrial partners is also under development. Another CRADA partner was introduced to Diffraction International last year and has indicated significant interest in mutual programs, some involving ORNL. In addition, an industrial partner that is developing new methods for manufacturing aspheric optical components has requested that ORNL provide on-process optical metrology development and has indicated interest in working with Diffraction International to develop CGHs for use in their optical metrology facility.

8. CONCLUSIONS

This has been by all standards an incredibly successful CRADA. It exemplifies the type of cooperative interaction between a national laboratory and an industrial partner that can lead to significant technical and commercial achievements that have technical and economic benefits well beyond those enjoyed by the CRADA partners.

Appendix A

Affordable Aspheric Testing Technology Comes to Light

Affordable Aspheric Testing Technology Comes to Light

L. Curt Maxey
Oak Ridge National Laboratory

Optical systems with the best performance in the smallest, lightest packages require the use of aspheric elements

Better, smaller, lighter are popular adjectives of progress that are pushing state-of-the-art optical systems away from the use of spherical surfaces. Optical systems use lenses and mirrors with spherical surfaces whenever possible, because spheres are relatively easy to manufacture and test. To get the best performance, using spherical surfaces alone, requires numerous optical elements. Using aspheric components significantly reduces the number of elements required. Optimum performance in the smallest, lightest package is only achieved by including aspheric components in system designs.

Throughout history, testing technology has limited the production of aspheric optics

Aspheric optical components were first used in astronomical telescopes. Early astronomers produced parabolic mirrors for their Newtonian telescopes by first rubbing together two round flat glass disks to produce spherical surfaces. The astronomers then meticulously hand-polished one of the resulting mirror blanks to change its shape from a sphere to a paraboloid. Changes to the mirror shape were made on the basis of the images of the stars seen through the telescope. The astronomers' capacity for interpreting the images of stars determined the quality of the mirrors they produced. Then, as now, testing technology posed the greatest limitation to the production of quality aspheric optics.

The Hubble Space Telescope misfortune graphically revealed the limitations of conventional tests based on null lenses

Modern optical designs contain aspheric surfaces with submicron tolerances. Although some of these optics are still produced by hand polishing, computer-controlled manufacturing is improving quality and speeding production. A few manufacturers have proprietary computer-controlled polishing processes for producing aspheric optics in glass and other brittle materials. Single point diamond turning is now routinely used to rapidly produce precision aspheric optics in metals, plastics, and selected crystalline materials.

Testing technology has not kept pace with the advances in aspheric manufacturing processes. Most precision optics are tested with laser interferometers that approach nanometer resolution but are specifically designed to measure spheres. To measure aspheres, special "null lenses" are constructed using spherical optics. Null lenses are designed to bend the light from the interferometer into a shape that matches the aspheric optic. The misfortunes of the Hubble Space Telescope graphically demonstrate the limitations of conventional null lenses.

The accuracy of a null lens is highly dependent on the spacing of individual lens elements. Each time a null lens is constructed by an optical metrologist, skill and attention become the only deterrents to disaster. If an error in assembly occurs, it cannot be detected without constructing another null lens to corroborate the first test. Even then, costly corroborative tests do not guarantee quality.

Whenever optical metrologists build null lenses, their skill and attention to detail are the only deterrents to disaster

In the case of the Hubble Space Telescope, the primary mirror was tested with two different null lenses. When the results did not agree, the lens with the more complex design was believed to be more accurate and was used for the final manufacturing processes. There was no way to know that the lens had been assembled with a small error in spacing. The result, now well known, was a large expensive mirror precisely manufactured to the wrong shape and deployed into space before the error was detected. Ultimately, a space mission was required to install expensive corrective (aspheric) optics.

The computer-generated hologram, or CGH, is a significant advance in aspheric testing and was used to test the corrective optics for the Hubble Space Telescope

Although it has only recently been commercialized, a significant advance in aspheric metrology was developed in the early '70s. Computer-aided design and computer-controlled manufacturing were combined to produce a new type of null lens known as a computer-generated hologram (CGH). Using diffraction, the phenomenon that produces the rainbow of color in a compact disk, the CGH bends light to match the shape of a specific aspheric optic.

Before being deployed in space, the corrective optics for the Hubble Space Telescope were certified using CGHs. A more assertive testimonial to the CGH technology would be difficult to conceive.

The advantage of using a computer-generated hologram for aspheric testing is that the design and manufacturing processes are highly repeatable. Once characterized, they can be trusted to produce reliable CGHs again and again. The repeatability of computerized design and manufacturing processes replaces the uncertainty inherent in the manual assembly and alignment of lens elements.

CGH's are produced with computerized reliability and repeatability

Despite their significant advantages, CGHs have remained a novelty item for the optical manufacturing industry as a whole. Designing or modifying optical interferometers to incorporate CGH correctors is difficult and expensive. A very few organizations have had the technical and financial wherewithal to successfully implement CGH testing. For the rest of the industry, however, the technology has been inaccessible. The design, production, and integration of CGH correctors requires uncommon levels of technical expertise and financial resolve.

Until now, few organizations have had the technical expertise and financial resolve to enjoy the advantages of CGH technology

True commercialization of CGH testing technology did not occur until 1992, when APA Optics marketed the Interferometer for Aspheric Testing. For the first time, optical test equipment was offered to industry with an integrated CGH interface. The product has not been a commercial success, despite strong industrial interest in the technology. The offering price of well over \$100,000 and the lack of clearly established technical success have significantly depressed the potential market.

The CGH Null Adapter places CGH technology within the reach of all aspheric optics manufacturers

The most recent commercialization of CGH testing technology offers availability, affordability, and compatibility. The CGH Null Adapter, recently marketed by Diffraction International, Ltd. places CGH technology within the reach of all aspheric optics manufacturers. The hardware is available at a very low cost and is designed for use with the existing test equipment common to most optics shops.

Solutions to the most common aspheric metrology problems are deliverable in less than a month for about \$5,000

Total hardware costs to test a typical asphere with the CGH Null Adapter are about \$5,000. Additional custom holograms for testing other conventional aspheres can be purchased for around \$3,500. Even exotic aspheric optics, with unusual shapes and steep surfaces, can be tested by using more elaborate CGH correctors. These are produced through more complex manufacturing processes and are available at about two to four times the cost of the standard holograms.

Custom CGHs are now more affordable and available than conventional null lenses. CGH technology eliminates the need for expensive and obscure glass types. Computerized control provides enhanced reliability and enables automated production. The tireless reliability of electron beam lithography replaces the labor-intensive grinding, polishing and alignment of conventional null lens elements. Custom CGHs are available from Diffraction International in two to four weeks.

For the first time in optics manufacturing history, solutions to the most common aspheric measurement problems are deliverable in less than a month for about \$5,000. The technology that ensured the successful correction of the Hubble Space Telescope is now within the reach of all optics manufacturers.

Appendix B

Cooperative Research Brings Advanced Optical Testing Technology to Market and to DOE Missions

Diffraction International
ER-LTA CRADA Hilite/Success Story

**Cooperative Research Brings Advanced Optical Testing Technology
to Market and to DOE Missions**

The technology that ensured the successful correction of the Hubble Space Telescope is now within the reach of all optics manufacturers. Diffraction International, a small Minnesota company, recently launched a successful commercial product with the help of an Oak Ridge National Laboratory (ORNL) Cooperative Research and Development Agreement (CRADA) funded by the DOE Energy Research Laboratory Technology Applications Program. Commercialization of this new aspheric testing device has resulted in a product that now provides almost 50% of Diffraction International's revenues.

Optical systems with aspheric components require fewer elements and achieve higher levels of performance than traditional spheric-based systems. Computational advances have simplified the tasks associated with designing these systems and computer controlled manufacturing has enhanced their producibility. Aspheric testing technology has not kept pace, however, and has represented one of the most expensive and least reliable links in the aspheric manufacturing process.

Diffraction International's new product, the Computer Generated Hologram (CGH) Null Adapter, is a simple accessory that allows accurate and less expensive testing of aspheric optical surfaces using standard commercially available test equipment. CGHs represent the most reliable technology for aspheric testing because they are produced with computer controlled reliability and repeatability throughout each step of their manufacture. Previously, the entry level cost of well over \$100,000 has kept CGH testing technology in the esoteric domain of a very few well-financed manufacturers. The CGH Null Adapter reduces the entry level cost for CGH testing to about \$5,000.

Multiple benefits to DOE and ORNL have accrued from the work performed under this CRADA. The direct experience gained in implementing CGH tests for high precision aspheric surfaces ensures that the capability has been successfully demonstrated and validated within the DOE/ORNL organization. This is essential to our ability to respond rapidly and realistically when called upon to assess technical feasibility and cost for programs incorporating advanced optical systems.

Equally important are the indirect benefits to any existing or emerging DOE programs that will require advanced optical systems. These systems have been and will continue to be of importance in applications as diverse as non-proliferation and arms control, directed energy for research applications and above ground testing, development of advanced materials and manufacturing processes, development of alternative energy sources, instruments for environmental research and systems for global monitoring.

The ability to manufacture the necessary aspheric optical components for these systems is directly limited by the ability to test them. The CGH Null Adapter provides the means for inexpensively incorporating the most reliable aspheric testing technology into the development these optical systems. For programs in which alternate testing technologies have already been selected, the

Appendix C

Figure Metrology of Deep General Aspherics Using a Conventional Interferometer with CGH Null

PROCEEDINGS REPRINT

SPIE—The International Society for Optical Engineering

Reprinted from

Optical Manufacturing and Testing

9–11 July 1995
San Diego, California

Volume 2536

©1995 by the Society of Photo-Optical Instrumentation Engineers
Box 10, Bellingham, Washington 98227 USA. Telephone 360/676-3290.

Figure Metrology of Deep General Aspherics Using a Conventional Interferometer with CGH Null

Steven M. Arnold

Diffraction International
11345 Highway 7, #421, Minneapolis, Minnesota 55305-5300

L. Curt Maxey, J. E. Rogers and R. C. Yoder

Martin Marietta Energy Systems, Inc. (ORNL)
P.O. Box 2008, Oak Ridge, Tennessee 37831-6085

ABSTRACT

We present a simple and general method of aspheric figure metrology using a CGH null mounted in the test beam of a conventional Fizeau or Twyman-Green interferometer. A "standard" reflective CGH is used to establish optical alignment with respect to the interferometer's spherical test beam. This alignment is then mechanically transferred to a custom CGH null. The accuracy of the alignment transfer is readily verified. The test method has been modeled by raytracing and verified experimentally by testing a perforated 8-inch F/1.5 on-axis paraboloid and a 50 mm F/2 off-axis paraboloid from their centers of curvature.

Keywords: optical testing, aspheric metrology, interferometry, CGH, computer generated holograms.

1. OVERVIEW OF TESTING METHOD

Interferometric null testing of aspheric optics is one of the most challenging tasks commonly encountered by an optical metrologist. Aspheric null tests require the use of auxiliary optics and the alignment requires precision adjustment in many degrees of freedom. Moreover, the alignment process usually requires simultaneous iterative alignment of the test optic and the auxiliary null optic to the interferometer. Systems which eliminate the iterative nature of the alignment process greatly simplify testing in specific applications¹.

Meticulous alignment of aspheric null tests is essential, for any residual alignment errors will be interpreted as errors in the test optic. In practice, the alignment process is considered complete when no further improvement in the test results can be obtained through alignment adjustments. The ultimate quality of the test results is influenced by the alignment skills of the metrologist.

1.1 CGH Null Adapter

The CGH Null Adapter, depicted in Figure 1, consists of a three-axis mount to which framed, prealigned CGH nulls are attached. A custom CGH null transforms the spherical or collimated test beam into a desired aspheric wavefront. Depending on available space, the CGH Null Adapter can be located on either the interferometer side or the test optic side of the CGH null. CGH frames are kinematically attached to the Null Adapter by magnetic coupling. Any single CGH frame may be easily removed and replaced to micron accuracy. One CGH may be replaced with another to an alignment accuracy of several microns.

Figure 1. CGH Null Adapter

When using the CGH Null Adapter for aspheric testing the iterative alignment problems inherent to most other null test methods are eliminated. This is possible because of the Alignment CGH and kinematic mount, which establish the alignment of the CGH Null to the interferometer. During the test setup, the Alignment CGH (which mimics a reflective sphere) is placed at a pre-determined distance from the interferometer focus. At this location, the tilt, tip and focus of the Alignment CGH are aligned (just as if it were a sphere) to produce a null interferogram.

Once the null interferogram is obtained, the Alignment CGH is removed from the kinematic mount and replaced with the CGH Null. Without further adjustment, the diffractive surface of the CGH Null produces an aspheric wavefront corresponding to the ideal shape of the aspheric test optic. The test optic is then aligned to this wavefront using a five-axis mount. Adjustments of tilt and centration are interactive and must be adjusted in concert. Initial alignment of the test optic to the aspheric test wavefront is simplified by the inclusion of fiducials in the design of the CGH null. These fiducials are visually aligned to the edges of the test optic, simplifying the centration adjustments.

1.2 Definition of coordinates and degrees of freedom

We adopt a right-handed coordinate system (Figure 2) with its origin at the center of the CGH null. The z axis is perpendicular to the CGH and directed toward the test optic. Depending on the particular test configuration, the focus (X_F, Y_F, Z_F) of the spherical test wavefront may be located before the CGH (as illustrated), between the CGH and the test optic, or beyond the test optic (convex asphere).

The global z axis intersects the test optic at a distance Z_T from the CGH face. A local coordinate system of the test optic is obtained by translating the global CGH coordinate system to the point $(0,0,Z_T)$. Unlike flats and spheres, aligning an aspheric optic to its null test wavefront requires a minimum of five degrees of freedom. We define decentration DCX and decentration DCY as translations along the x and y axes respectively. We define focus adjustment DCZ as translation along the z axis. Rotations TLA and TLB about the x and y axes respectively are defined as pitch and yaw (regardless of which direction is vertical).

In the case of an off-axis aspheric segment, the z-axis intersection point $(0,0,Z_T)$ may be near the physical center of the test optic rather than at its optical vertex. This is accommodated by specifying non-zero design values for some of the test optic degrees of freedom, typically DCY, TLA and DCZ. A sixth degree of freedom, clock rotation TLC about the test axis, is required if the off-axis segment is to be physically centered as well as optically nulled. In practice, off-axis aspheric segments are frequently tested in five-axis mounts by using alignment fiducials or mechanical methods² to fix the clock angle. Both of the off-axis segments discussed here were tested in this way.

Figure 2. CGH Null Test Coordinate System

1.3 CGH Null Design

In designing a custom CGH null for use with the CGH Null Adapter, we begin by tracing the null wavefront from the aspheric test optic. This is most easily accomplished by modeling the test optic as a refractive surface and setting the object space refractive index to zero. It is best to start at the test optic even if the optical design software does not allow a refractive index of exactly zero because the CGH phase function will likely require high order terms which make the test aperture difficult to control when tracing from the interferometer side.

In tracing from the test optic, the null wavefront will come to a best focus at some point—virtual in the case of a convex asphere. The CGH null is placed at a distance Z_F from this best focus which matches one of the 6 positive and 6 negative “radii” of our standard Alignment CGHs. The choice is often limited by the need to keep the CGH on the proper side of the test optic, out of the region of the caustic (where rays of the null wavefront intersect), and outside the interferometer optics.

Our next step is to design a CGH null which will diffract the null wavefront to a perfect focus at the on-axis point $(0,0,Z_F)$. If possible, we prefer that the test optic and CGH null have the same optic axis since this will simplify the design. The CGH phase function is usually placed on the surface toward the interferometer since this is the orientation in which the Alignment CGH works. For afocal test configurations ($Z_F = \infty$), we instead reverse the CGH so that its substrate thickness and index will not enter into the design.

Finally, we add a carrier to the CGH null so that the various diffraction orders will be separated. This carrier is a difference between two spherical waves: one at the design focus $(0,0,Z_F)$ and the other at the cat's eye focus $(X_F,0,Z_F)$. Most optical design software can best model this type of carrier as an HOE on a dummy surface coincident with the CGH phase function. The Alignment CGH determines the choices of spatial relationship between the design focus and the cat's eye focus. We have produced three different alignment CGHs for which the foci are 0, 1 and 2 degrees from the defined CGH axis. The 0 degree Alignment CGH is suitable for perforated on-axis aspheres which can be refocused to introduce an axial carrier. Other aspheres use the 1 or 2 degree alignment CGHs depending on the maximum aspheric slope. Note that by placing the CGH surface toward the interferometer, we can usually avoid tracing a non-axially-symmetrical wavefront through the substrate.

After designing a CGH null in single pass, we analyze the test configuration in double pass beginning (and ending) with a point object/image at the cat's eye position $(X_F,0,Z_F)$. In particular, we look at the effects of CGH misalignment (displace object/image from X_F, Y_F, Z_F), asphere misalignment (DCX, DCY, DCZ, TLA, TLB, TLC) and unwanted diffraction orders.

2. EXPERIMENTAL RESULTS

Testing was conducted using two interferometers, one a Twyman-Green configuration manufactured by Breault Research Organization (BRO VIP) and the other a Fizeau type manufactured by Zygo (Zygo Mark IV).

Three paraboloidal mirrors were tested with the CGH Null Adapter. The selected optics (two off-axis segments and one on-axis paraboloid with a central obscuration) afforded good variety in their dimensional parameters and overall quality. Paraboloids were selected because they are easily tested in an autocollimating configuration using a reference flat as the requisite auxiliary optic³. Test results obtained in the autocollimating configurations were compared with those obtained using the CGH nulls.

The CGH nulls used in these tests were fabricated as chrome on glass patterns, diffracting approximately 10 percent into the first order and 25 percent into the zeroth order. Since the paraboloids were all of high reflectivity, nearly 1 percent of the test beam was returned to the interferometer, yielding good fringe visibility.

2.1 Alignment CGH

As a first step in each test setup, an Alignment CGH is placed at a pre-determined distance from the interferometer focus. At this location, the tilt, tip and focus of the Alignment CGH are aligned (just as if it were a sphere) to produce a null interferogram.

Two different Alignment CGHs were used during the testing procedures. The first was designed as an F/4 first order zone plate and was manufactured using low-reflectivity chrome. Its aperture was spatially partitioned into eighteen pie-shaped segments: three each of focal radii 20 mm, 33 mm, 55 mm, 90 mm, 150 mm and 250 mm. These radii were selected to accommodate the back focal distances of common Fizeau spheres and to avoid confusion amongst higher diffraction orders. The zones for each focal radius were decentered in the x direction to yield a 1 degree angle between the CGH axis and the respective focal point. The initial results with this first order Alignment CGH showed that the alignment procedure

was simple and intuitively obvious for all of the different radii. By observing the nulled interferogram (Figure 3), it was demonstrated that the CGH could easily be removed and replaced to micron tolerances within the CGH Null Adapter.

The second Alignment CGH was designed as an $F/2$ zone plate to provide greater alignment sensitivity. The design was modified to use the third diffracted order so that the feature sizes at the decreased $F/\#$ would remain reasonable. The diffraction pattern was fabricated in bright chrome to recover the lost diffraction efficiency and a multilayer anti-reflective coating was added to the rear surface of the optic.

Alignment of this third-order Alignment CGH was complicated by the dominant first order diffraction. Intuitively, the CGH was aligned to the (brighter) first order, rather than the third, resulting in incorrect positioning. Fortunately the error was easily detected because of a large amount of spherical aberration in the resulting interferogram. By carefully re-positioning the Alignment CGH at the specified distance from the interferometer focus, correct results were easily obtained.

Figure 3. Alignment CGH Interferogram

2.2 Off-Axis $F/2$ Paraboloid (A8037255)

The first optic tested with the CGH Null Adapter was an off-the-shelf off-axis paraboloid (Janos P/N A8307255). This is a 50.8 mm diameter aperture, diamond-turned segment with a 101.6 mm focal length, centered 15 degrees off axis.

The 40 mm aperture CGH null is depicted in Figure 4 (each fringe representing 25 fringes). The paraboloid axis is tilted 15 degrees and optically decentered in the +Y direction to bring its geometric center into coincidence with the CGH axis. The interferometer focus ($F/3.5$) is located at $X_F=2.618$ mm, $Z_F=150$ mm with respect to the CGH. This 1 degree carrier is perpendicular to the symmetry axis of the paraboloid, so the CGH null has no symmetry. The required aspheric correction is about 200 waves. The CGH null includes four small fiducials placed just outside the test optic aperture but within the CGH aperture.

This optic was first tested using the Twyman-Green interferometer. After aligning the CGH Null Adapter and inserting the CGH null, the test optic (in a five axis mount) was placed in the diffracted aspheric test wavefront. No attempt was made to accurately center the optic using the alignment fiducials provided in the CGH. The test optic was positioned to refocus the interferometer test beam to a point coincident with the focus of the diverger. Because of the tilt carrier in the CGH null design, a series of diffracted spots was obtained. The undiffracted zero order was easily identified by removing the CGH null. Upon replacing the CGH null the other spots were identified relative to the zero order.

Some initial confusion resulted from the fact that the first spot adjacent to zero order was not useful for testing. The second spot contained less aberration and produced useful results. Ultimately, it was recognized that the CGH null produced diffracted wavefronts between the expected zeroth and first diffracted orders. This occurs because the CGH null works in double pass, diffracting the test beam twice (first: first). The first (and brightest) spot is a combination of the zeroth: first and first: zeroth orders which, together, are five times brighter than the desired first: first spot. With this understanding the useful order was recognized as the second returned spot (counting from the zero order). We later modified our raytrace model to plot the expected spot patterns (Figure 5).

After identifying the correct diffracted order, the alignment mode of the interferometer was used to focus the returned light onto the CCD array. The returned light from the reference arm

Figure 4. Off-Axis CGH Null A8037255

of the interferometer was temporarily moved off axis. In this configuration, it was possible to use a "star test"⁴ method to analyze the returned light from the test optic. By reducing the intensity of the test beam, the star test provided very useful alignment information. Using a defocused spot, the y-translation and pitch adjustments were simultaneously adjusted to maximize the bilateral symmetry. Next the x-translation and yaw were adjusted to maximize the rotational symmetry and minimize coma. Focus was then adjusted to produce the minimum spot size. Following the star test alignment, the interferometer was returned to its normal operating configuration and a good interferogram was obtained. Minimal additional adjustments were made to achieve the best overall interferogram quality (Figure 6).

Figure 5. Simulated CGH Star Test of A8037255

After tests were completed with the Twyman-Green interferometer, the CGH Null Adapter and the test optic were moved to the Fizeau interferometer for comparative tests. Having arrived at a useful alignment procedure based on the star test, it was disappointing to learn that, with the Fizeau interferometer in alignment mode, the intensity could not be sufficiently reduced to obtain a useful star test.

Faced with this equipment difference, the CGH alignment fiducials were used to position the test optic as precisely as possible. Four radial fiducials placed just outside the test optic aperture were used to center the test optic while focus was adjusted to maintain best focus in the returned light. Here the tilt carrier in the CGH produced an unanticipated result. The fiducials at the $\pm Y$ edges of the test optic (perpendicular to the carrier tilt) were most useful because the multiple images corresponding to multiple diffraction orders were distributed parallel to the test optic edge. The $\pm X$ fiducials were less useful because the orders overlapped on the test optic and selecting the most prominent zero order resulted in a biased centration alignment. In retrospect, it may have been better to place the fiducials to use the bright zero order images.

After obtaining the best possible alignment using the CGH fiducials, the interferometer was returned to test mode and a useful interferogram was obtained. Observing that the interferogram was not bilaterally symmetric, x-translation and yaw were carefully adjusted in an iterative process to maximize this symmetry. Next, y-translation, pitch and focus were iteratively adjusted to reduce the residual astigmatism and improve overall interferogram quality.

Figure 6. CGH Interferogram of A8037255

Figure 7. Autocollimation Test of A8037255

Using the Twyman-Green interferometer, an autocollimating test (Figure 7, above) produced interferograms virtually identical to the CGH null test; the autocollimating test was double pass, of course, approximately doubling the fringe density. The distinguishing feature is a 3-lobed pattern, likely due to the three tapped mounting holes. Quantitative comparison of the test results requires compensating the autocollimating test for the cosine error resulting from the 15 degree off-axis angle. The resulting peak-to-valley measurements are 3.68 waves for the CGH test and 3.79 waves for the autocollimation test—agreement to within approximately 0.10 wave despite the large magnitude of the figure error. Comparison of rms results after compensating for cosine error reveals 0.691 waves rms for the CGH test and 0.707 waves rms for the autocollimating test—agreement better than 1/60 wave.

The autocollimating test results obtained with the Fizeau interferometer indicate 3.49 waves peak-to-valley and 0.743 waves rms. The CGH test with the same instrument indicates 3.60 waves peak-to-valley and 0.679 waves rms. Agreement between the two tests with the Fizeau interferometer is approximately 0.10 waves peak to valley and 1/15 wave rms.

2.3 Perforated F/1.6 Paraboloid (ORNL8IN)

This second optic tested was a diamond-turned, aluminum, on-axis paraboloid with a focal length of 300 mm and an aperture of approximately 180 mm diameter. The optic contained a central hole of approximately 50 mm diameter.

The test configuration uses the on-axis Alignment CGH and on-axis CGH null (Figure 8, also with 1 fringe representing 25 fringes). The CGH null is located 90 mm after the focus of the interferometer and has about 250 waves of negative focal power to serve as an on-axis carrier. The test optic is located 505 mm beyond the CGH face.

In order to test this optic in autocollimation, it was necessary to use a perforated optical flat. An 200 mm diameter, diamond-turned, aluminum flat was used for initial testing performed with the Twyman-Green interferometer. This flat was believed to have errors on the order of those measured in the paraboloid, however, so a better perforated flat was sought for the remainder of the testing. The largest high-quality perforated flat that could be located was 150 mm in diameter, limiting the autocollimation test to F/2. During both the autocollimating test and the CGH tests, a mask was placed over the optic so that the same 150 mm portion of the optic was tested. Due to the limited time available with the 150 mm flat, an autocollimating test was performed only with the Fizeau interferometer.

Initial CGH test results with the Twyman-Green interferometer were not satisfactory, as a clear interferogram could not be obtained. The Fizeau interferometer however produced excellent interferograms with the CGH test configuration (Figure 9). It was ultimately determined that the greater field of view of the Twyman-Green interferometer was allowing extraneous

Figure 8. On-Axis CGH Null ORNL8IN

Figure 9. Fizeau CGH Interferogram of ORNL8IN

orders to enter the viewing arm, reducing contrast. Careful spatial filtering with an iris diaphragm (reduced to approximately 0.5 mm diameter) at the focus of the diverger eliminated the unwanted orders and resulted in excellent interferogram quality. A raytrace model plot (Figure 10) confirms that the 5 mm defocus was only marginally adequate.

The autocollimating test of this paraboloid with the Fizeau interferometer revealed a peak-to-valley error of 0.362 waves and an rms error of 0.064 waves. The CGH test with the same interferometer revealed an error of 0.361 waves peak to valley and 0.067 waves rms. The CGH test of this optic with the Twyman-Green interferometer revealed an error of 0.372 waves, in good agreement with the value obtained with the Fizeau; the rms error was 0.080, slightly higher than that obtained with the Fizeau.

2.4 Rectangular F/1.4 Paraboloid (ORNLASMT)

The final optic tested was a rectangular aperture off-axis paraboloid with a 300 mm focus. The long edge of the 200 mm by 87.5 mm aperture is located approximately 25 mm from the vertex of the parent paraboloid. This mirror was machined into a beryllium substrate and hand polished to final figure, resulting in asymmetric and complex residual errors.

The CGH null (Figure 11) has an aperture of 34 mm x 18 mm. The F/3 interferometer focus is located at $X_F=1.571$ mm and $Z_F=90$ mm with respect to the CGH, resulting in a 1 degree carrier. The test optic was tilted and decentered to bring its geometric center into coincidence with the CGH axis. The total aspheric correction is about 275 waves. The 1 degree carrier is perpendicular to the test optic symmetry axis so the CGH null has no symmetry.

This optic was by far the most challenging of the three test optics to align to the CGH null. This was due to the combination of large complex aberrations in the optic and the lack of alignment fiducials in the CGH. Alignment using the Twyman-Green interferometer was achieved by using the star test method described previously. A raytrace simulation is shown as Figure 12. Careful systematic application of the star test method was sufficient to achieve a very high quality of alignment. Final alignment adjustments after returning the interferometer to view mode required only very small adjustments of tilt and focus. The interferogram is shown as Figure 13.

Figure 11. Off-Axis CGH Null ORNLASMT

Figure 10. Simulated CGH Star Test of ORNL8IN

Figure 12. Simulated CGH Star Test of ORNLASMT

Alignment to the Fizeau interferometer was much more challenging, due to the fact that the star test method could not be successfully implemented with this interferometer. Initial alignment was achieved through interpretation of the image and then the interferogram, making adjustments to maximize the symmetry in each. Final alignment was achieved through meticulous interferogram interpretation and alignment corrections.

The high slope errors at the edges of the optic made the numerical results highly sensitive to small variations in aperture placement when processing the data. Using the Fizeau interferometer for the autocollimating test, the peak-to-valley error was measured as 1.9643 waves and the rms error was measured as 0.330 waves. The CGH data with the same interferometer yielded 1.81 waves peak-to-valley and 0.394 waves rms.

This mirror had been tested previously with the Twyman-Green interferometer in the autocollimating test and the file data (Figure 14) was used for comparison with the CGH test data. In both cases the optic was tested to the maximum extent possible with the interferometer. The very high slope due to rolloff induced during the hand polishing produced large errors at the edge of the data set. As a result, both tests revealed peak to valley errors about 2.5 waves. The rms errors were given as 0.312 waves for the autocollimating test and 0.357 for the CGH test. Excluding a very thin margin around the edge of the data set, well over 95% of the data obtained in both Twyman-Green tests falls in the range of 1.8 to 1.9 waves.

The wider field of view of the Twyman-Green interferometer is believed to be responsible for the much larger peak-to-valley errors in the raw data. The Fizeau interferometer appears to be rejecting the light from the high slope regions of the optic resulting in significantly lower peak-to-valley figure errors. Accurate comparison of the test results would require physically masking the optic to ensure that the same portions of the aperture are being measured by both interferometers.

Figure 13. CGH Interferogram of ORNLASMT

Figure 14. Autocollimation Test of ORNLASMT

3. SENSITIVITY AND TOLERANCE ANALYSIS

The method we have described for testing of aspheric optics with CGH nulls has an accuracy which is dependent on several possible error sources. These can be readily quantified to produce a useful and verifiable estimate of test accuracy. This is in contrast to other CGH test methods which require a detailed knowledge of the interferometer optics⁵.

3.1 Error Sources

Error sources can be classified as CGH pattern errors and other errors. CGH pattern errors include encoding and digitization errors and errors in writing a CGH master. CGH pattern errors may also include contact printing errors (for duplicated masters) and etching errors (for phase type CGHs). Verification and certification of CGH pattern errors is the subject of a companion paper in this volume⁶.

In the cases presented, the major error sources are interferometer errors such as the reference mirror or Fizeau sphere and the transmitted wavefront distortion (TWD) of the CGH substrate. To control the later, we print the CGH nulls onto precision photomask substrates which have a TWD of nominally 0.10 wave or better. For chrome CGHs, it is possible to measure the substrate TWD in zero order and subtract it from test results. This was not attempted in the current work.

It is also possible that the aspheres were tested assuming an incorrect radius of curvature. This seems unlikely since the primary residual aberration is not spherical. It is of course possible to verify radius of curvature to high accuracy by measuring the distance along the test axis from cat's eye focus to the test optic with a radius slide. This feature was not available. The CGH substrate thickness is another possible source of spherical error, particularly for faster aperture tests. Fortunately, the index of silica is well known and the substrate thickness can be directly measured.

Using the CGH Null Adapter, alignment and framing errors are well controlled and readily verified. We encounter variable errors of 3 degrees of freedom in optically aligning the Alignment CGH to the interferometer test beam. We also have fixed errors of 6 degrees of freedom in prealigning each CGH in its frame. This all reduces to just the 3 translational degrees of freedom which define where the cat's eye focus (X_F, Y_F, Z_F) is located relative to the CGH null. The tilt, tip and focus fringes of the alignment interferogram and the CGH framing errors are all simply related by their effect on cat's eye focus location. The absolute magnitude of CGH framing errors is somewhat difficult to define and measure, but the relative errors between an alignment CGH and a CGH null can be determined with no special metrology equipment other than the Fizeau or Twyman-Green interferometer, the CGH Null Adapter, and a microscope with a crosshair reticle and 20X objective. Relative alignment error is minimized by attaching the CGHs to the Null Adapter in their preferred or keyed orientation, matching that used during prealignment.

Because it acts like a return sphere, the Alignment CGH can be readily verified using an interferometer with radius slide. Each zone plate aperture works at both positive and negative focal radius and the CGH pattern error reverses sign while the substrate flatness error does not. Consequently, two interferograms (with the CGH clocked 180 degrees) are sufficient to isolate CGH pattern error.

3.2 Test Configuration Analyses

The following tables summarize the sensitivity and tolerance analysis results for the three test configurations described. The sensitivities to misalignments, laser wavelength, and CGH substrate thickness and refractive index were determined from a raytrace model of each test configuration. The sensitivity to registration and printing errors is based on the CGH minimum grating period. Other sensitivities are for the most part obvious. These results are consistent with the observed agreement between CGH null test interferograms and autocollimation interferograms.

Janos A8037-255 Off-Axis Paraboloid				Tolerance Analysis of CGH Null Test				
Surface or Group	Raytrace Model		Design Value	Design Increment	Sensitivity/Tolerance Analysis			
	Parameter	Unit			P-V Waves	Budgeted Tolerance	P-V Waves	P-V Percent
All	Baseline Model P-V	fringe	0.00097	1	1.00000	0.00097	0.001	0.4%
All	Wavelength	um	0.63282	0.0001	0.00098	0.00001	0.000	0.0%
0	Fizeau Sphere	wv	0	1	1.00000	0.05	0.050	22.1%
0	Decenter Fizeau (X)	mm	0.698	0.01	0.00401	0.0005	0.000	0.1%
0	Decenter Fizeau (Y)	mm	0.000	0.01	0.00320	0.0005	0.000	0.1%
0-1	Focus Fizeau (-Z)	mm	-150.000	0.01	0.01195	0.0116	0.014	6.1%
1-2	Transfer X	mm	0.000	0.01	0.00201	0.010	0.002	0.9%
1-2	Transfer Y	mm	0.000	0.01	0.00160	0.010	0.002	0.7%
1-2	Transfer Z	mm	0.000	0.01	0.00597	0.010	0.006	2.6%
1 CGH	Encoding	wv	0	1	1.00000	0.020	0.020	8.8%
1 CGH	Digitization	wv	0	1	1.00000	0.018	0.018	8.0%
1 CGH	Registration	um	0	1	0.01912	0.300	0.006	2.5%
2-3 CGH	Contact Printing	um	0	1	0.01912	0.400	0.008	3.4%
1-2 CGH	Thickness	mm	2.327	0.1	0.00225	?	0.000	0.0%
1-2 CGH	Index Silica		1.45702	0.001	0.00156	0.0001	0.000	0.1%
1-2 CGH	TWD	wv	0	1	1.00000	0.1	0.100	44.2%
3	Vertex Radius	mm	-203.2	1.00%	0.56313	N/A	0.000	0.0%
3	Conic Constant		-1	0.01	0.67318	N/A	0.000	0.0%
Sum							0.226 wv	
RSS							0.116 wv	

ORNL 8-inch Paraboloid
S/N 941010B

Tolerance Analysis of CGH Null Test

Surface or Group	Raytrace Model			Sensitivity Analysis		Tolerance Analysis		
	Parameter	Unit	Design Value	Design Increment	P-V Waves	Budgeted Tolerance	P-V Waves	P-V Percent
All	Baseline Model P-V	wv	0.00006	1	1.00000	6E-05	0.000	0.0%
All	Wavelength	um	0.63282	0.0001	0.00006	0.00001	0.000	0.0%
0	Fizeau Sphere	wv	0	1	1.00000	0.050	0.050	23.3%
0	Decenter Fizeau (X)	mm	0.000	0.01	0.00349	0.000	0.000	0.1%
0	Decenter Fizeau (Y)	mm	0.000	0.01	0.00349	0.000	0.000	0.1%
0-1	Focus Fizeau (Z)	mm	-90.000	0.01	0.01253	0.005	0.007	3.1%
1-2	Transfer X	mm	0.000	0.01	0.00349	0.010	0.003	1.6%
1-2	Transfer Y	mm	0.000	0.01	0.00349	0.010	0.003	1.6%
1-2	Transfer Z	mm	0.000	0.01	0.01253	0.010	0.013	5.8%
1 CGH	Encoding	wv	0	1	1.00000	0.022	0.022	10.2%
1 CGH	Digitization	wv	0	1	1.00000	0.010	0.010	4.7%
1 CGH	Registration	um	0	1	0.08190	0.300	0.025	11.4%
2-3 CGH	Contact Printing	um	0	1	0.08190	0.400	0.033	15.2%
1-2 CGH	Thickness	mm	2.344	-0.057	0.00085	-0.057	-0.001	-0.4%
1-2 CGH	Index Silica		1.45702	0.001	0.00006	0.0001	0.000	0.0%
1-2 CGH	TWD	wv	0	1	1.00000	0.050	0.050	23.3%
3	Vertex Radius	inch	-600.000	1.00%	0.24327	N/A	0.000	0.0%
3	Conic Constant		-1	0.01	0.26560	N/A	0.000	0.0%
							Sum	0.215 wv
							RSS	0.087 wv

ORNLASMT Paraboloid
Job Number C9403

Tolerance Analysis of CGH Null Test
Release Number 950302A

Surface or Group	Raytrace Model			Sensitivity Analysis		Tolerance Analysis		
	Parameter	Unit	Design Value	Design Increment	P-V Waves	Budgeted Tolerance	P-V Waves	P-V Percent
All	Baseline Model P-V	wv	0.00218	1	1.00000	0.00218	0.002	0.8%
All	Wavelength	um	0.63282	0.0001	0.02581	0.00001	0.003	0.9%
0	Fizeau Sphere	wv	0	1	1.00000	0.05	0.050	17.9%
0	Decenter Fizeau (X)	mm	1.571	0.1	0.08583	0.000	0.000	0.1%
0	Decenter Fizeau (Y)	mm	0.000	0.1	0.06100	0.000	0.000	0.1%
0-1	Focus Fizeau (Z)	mm	-90.000	0.1	0.25772	0.0053	0.014	4.9%
1-4	Transfer X	mm	0.000	0.1	0.08583	0.010	0.009	3.1%
1-4	Transfer Y	mm	0.000	0.1	0.06100	0.010	0.006	2.2%
1-4	Transfer Z	mm	0.000	0.1	0.25772	0.010	0.026	9.2%
2-3 CGH	Encoding	wv	0	1	1.00000	0.021	0.021	7.5%
2-3 CGH	Digitization	wv	0	1	1.00000	0.014	0.014	5.0%
2-3 CGH	Registration	um	0	1	0.04655	0.300	0.014	5.0%
2-3 CGH	Contact Printing	um	0	1	0.04655	0.400	0.019	6.7%
3-4 CGH	Thickness	mm	2.294	0.1	0.00395	0.005	0.000	0.1%
3-4 CGH	Index Silica		1.45702	0.0001	0.00219	0.0001	0.002	0.8%
1-4 CGH	TWD	wv	0	1	1.00000	0.100	0.100	35.8%
1-3 CGH	Phase Etch Uniformity	um	0.69233	1.00%	0.00250	N/A	0.000	0.0%
7	Vertex Radius	mm	-600.000	0.10%	0.05584	N/A	0.000	0.0%
7	Conic Constant		-1.000	0.0001	0.55817	N/A	0.000	0.0%
							Sum	0.279 wv
							RSS	0.121 wv

4. CONCLUSIONS

We have shown that CGH null testing of paraboloids using commercial interferometers can be easy to perform and accurate to a fraction of a wave peak-to-valley. The key is in kinematic fixturing and standardization of the test configurations to use an innovative Alignment CGH. Our results are in no way specific to paraboloids and will be equally true of more general aspheres. The paraboloids we tested were not of sufficient quality to demonstrate the ultimate limitations of the CGH null testing method. Our sensitivity and tolerance analyses suggest that, with software compensation of CGH substrate errors, the method should be accurate to 0.05 wave peak-to-valley in many cases.

5. ACKNOWLEDGEMENTS

This research was performed under a Cooperative Research and Development Agreement (CRADA No. ORNL93-0218) between Martin Marietta Energy Systems, Inc. and Diffraction International Ltd.

6. REFERENCES

-
- ¹ L. C. Maxey, "Novel Technique for Aligning Paraboloids," Proc. SPIE 1531, 147-155, 1991.
 - ² C. Bond and C. A. Pipan, "How to Align an Off-Axis Parabolic Mirror," Proc. SPIE 1113, 236-248, 1989.
 - ³ Daniel Malacara, *Optical Shop Testing*, pp. 755-757, John Wiley and Sons, New York, 1992.
 - ⁴ *ibid.* pp. 397-426
 - ⁵ S. M. Arnold, "An interferometer for aspheric testing: calibration and error compensation," ASPE Spring Topical Meeting, Tucson, April 7-9, 1992.
 - ⁶ S. M. Arnold and R. Kestner, "Verification and Certification of Aspheric Nulls," Proc. SPIE 2536, ?-?, 1989.

Appendix D

Diffraction International, Ltd. Product Literature and Advertisement Copy

Diffraction International

CGH & BINARY OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

CGH Null Adapter

At Diffraction International, we specialize in diffractive optics. Our staff has over 15 years experience in CGH and binary optics technology. As a small business, we offer attention, flexibility and fast response. Our goal is to become your single source of responsibility for design, engineering and fabrication of precision diffractive optics.

The CGH Null Adapter

Our CGH Null Adapter mounts on your optical table in the test beam of your Fizeau or Twyman-Green interferometer. A custom CGH null transforms the spherical or collimated test beam into the required aspheric wavefront.

Our CGH Null Adapter uses prealigned CGH frames and kinematic mounts. A special alignment CGH aligns the CGH Null Adapter to your interferometer.

Easy to Use

Step 1:

Position Null Adapter with attached alignment CGH in your interferometer's test beam.

Step 2:

Adjust tilt-tip (or decenter) and focus to acquire and null alignment fringes.

Step 3:

Remove alignment CGH and replace with custom CGH null.

Step 4:

Align test optic to acquire and null fringe pattern. Record interferogram data.

Step 5 (optional):

Subtract interferometer and CGH substrate errors (typically $\sim \lambda/10$).

Breaking the Rules

Experience says that a CGH null should be used in single pass because of low diffraction efficiency. It should be located in a collimated beam to ease alignment tolerances. Test and reference beams should pass similarly through the CGH to cancel any substrate aberrations.

Our CGH Null Adapter violates these rules, so we must produce high efficiency CGHs on precision substrates and align them to micron tolerances. You benefit in simple test configurations that use your existing interferometer and accessories.

Accurate

Our CGH nulls are encoded to meet your accuracy requirements. Substrate transmitted wavefront distortion, which is less than $\lambda/10$, can be measured in zero order transmission and subtracted.

System errors of your interferometer and transmission sphere can be compensated

in the usual manner — by measuring a good spherical surface and subtracting the result from subsequent interferograms.

Cost Effective

Custom nulls for our CGH Null Adapter are available in about 4 to 6 weeks. The cost will depend on your test requirements, but is generally less than for a conventional null.

Custom CGH Nulls

Designing a custom CGH null does not require raytracing of your interferometer or transmission sphere. The raytrace model begins and ends at a point source concentric with your transmission sphere.

We attempt to locate the CGH null at one of the twelve distances from focus accommodated by our standard alignment CGHs. The CGH must be outside the caustic region where rays cross. We make the CGH aperture somewhat oversized so that your test optic is the only

CGH Null Adapter

stop in the system. We favor a CGH aperture of 30 to 45 mm as a reasonable compromise between fabrication costs and alignment sensitivity.

Cylindrical CGH Nulls

We produce cylindrical nulls as fast as F/1.5. These are used with a transmission flat and incorporate their own alignment gratings. Our cylindrical nulls can be switched from positive to negative focal length by rotating the CGH frame 180° in its mount to reverse the tilt.

Full Documentation

Every CGH Null is shipped with full documentation. We provide an optical prescription of the CGH phase function and test configuration. We include a sensitivity and tolerance analysis which includes all CGH encoding and fabrication errors plus the alignment errors.

Interferogram Visibility

A CGH null reduces the intensity of the test beam, thus altering fringe visibility. In some cases, this may result in improved visibility, but in other cases visibility is degraded. Diffraction efficiency is 10 percent for chrome CGHs and 35 to 40 percent for binary CGHs.

Generally, we recommend binary phase CGHs for testing of uncoated glass surfaces and chrome CGHs for testing of mirror surfaces. Unless your test optic is anti-reflection coated, fringe visibility should be adequate.

Ghosts and Other Artifacts

Unwanted diffraction orders or reflections off the CGH substrate may create ghost fringes. These can usually be minimized or eliminated by careful test configuration design.

Companion Refractive Nulls

For testing of very deep or very fast aspheres, we recommend supplementing a CGH null with a simple refractive null. A precision singlet is usually sufficient to reduce the CGH bandwidth and alignment sensitivity and can be incorporated into the CGH mounting plate on a custom basis. Because we prealign the refractive null to the CGH, your alignment procedure remains simple.

A refractive null is most easily added to our Model 1206R reducer plate. Often, one refractive null can be used in hybrid null testing of several similar asphere designs.

Meeting Your Needs

We continually seek to expand our capabilities in response to customer needs. Please allow us the opportunity to address your specific aspheric testing requirements.

Additional Literature

- CGH Frames and Mounts
- HA50 Series Alignment CGHs
- CGH Cylinder Nulls
- CGH Photomasks
- Binary Optics

K. Creath and J. C. Wyant, "Holographic and Speckle Tests," chapter 15 of *Optical Shop Testing*, D. Malacara, Editor (Wiley, New York, 1992).

STANDARD CGH NULL TEST

CYLINDER NULL TEST

ALTERNATE CGH NULL TEST

HYBRID NULL TEST

Diffraction International

CGH & BINARY OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

HA50-Series Alignment CGHs

At Diffraction International, we specialize in diffractive optics. Our staff has over 15 years experience with aspheric CGH nulls. Our goal is to become your single source of responsibility for design and fabrication of diffractive null optics.

Alignment CGHs

The HA50 series alignment CGHs are used to optically align our CGH Null Adapter to your interferometer test beam and radius slide axis. On replacing the alignment CGH with a custom CGH null, the spherical or collimated test beam is diffracted to produce a prescribed aspheric wavefront.

How They Work

The alignment CGH is a reflective zone plate which diffracts your spherical test beam back onto itself to create an interferogram. By turning adjustment screws to remove tilt and focus fringes, the CGH

PATENT PENDING

coordinate system is optically aligned to the focus of your interferometer's spherical test beam.

Each alignment CGH has multiple pie-shaped apertures and acts like a multi-radius return sphere. By using plus and minus diffracted orders, the same alignment CGH works for both real and virtual foci.

Our 3-axis Null Adapter can be aligned to a spherical test beam. With our 6-axis Null Adapters you can additionally align the CGH axis parallel to your radius slide or else center and orient the CGH in your interferometer's field of view.

Tilt or Decenter?

For off-axis CGHs, one must choose between tilting or decentering. Tilting leaves the CGH centered in the field of view, but results in a test axis which is not parallel to the interferometer axis. Decentering keeps the test axis parallel to the interferometer (and radius slide) axis, but may require a faster transmission sphere to fill the decentered test pupil.

Three Models Available

We currently offer three standard alignment CGHs for use with our CGH Null Adapters and your 633 nm wavelength interferometer. Model HA50T0 provides 12 on-axis foci. Models HA50T1 and HA50T2 offer the same focal distances, but with 1 and 2 degrees tilt between interferometer and CGH axes.

Specifications

The table below gives CGH coordinates (z,x) of the 12 transmission sphere focus locations. Models HA50T1 and HA50T2 also include linear grating apertures for the purpose of aligning to a collimated test beam.

Meeting Your Needs

We can provide custom alignment CGHs for other radii or wavelengths. Please allow us the opportunity to address your aspheric testing requirements.

Easy Optical Alignment

Step 1:

Mount HA50 series alignment CGH facing interferometer and with the transmission sphere focus near prescribed CGH coordinates (z,x).

Step 2 (6-axis option):

Using a transmission flat, adjust tilt-tip to acquire and null tilt fringes. Replace flat with transmission sphere.

Step 3:

Adjust tilt-tip (or decenter) and focus to acquire and null alignment fringes.

Step 4:

Replace alignment CGH with custom CGH null.

Focus Distance Z	Focus Decenter: $X = Z \times \tan(\text{tilt})$			Maximum Null Aperture	
	HA50T0 (on-axis)	HA50T1 (1 deg tilt)	HA50T2 (2 deg tilt)	Model 50 38 x 38 mm	Model 50R Ø 48 mm
±20 mm	0.0 mm	0.349 mm	0.698 mm	F/0.57	F/0.42
±33 mm	0.0 mm	0.576 mm	1.152 mm	F/0.94	F/0.69
±55 mm	0.0 mm	0.960 mm	1.921 mm	F/1.57	F/1.15
±90 mm	0.0 mm	1.571 mm	3.143 mm	F/2.57	F/1.88
±150 mm	0.0 mm	2.618 mm	5.238 mm	F/4.27	F/3.13
±250 mm	0.0 mm	4.364 mm	8.730 mm	F/7.14	F/5.21

Diffraction International

CGH & BINARY OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

CGH Cylinder Nulls

At Diffraction International, we specialize in diffractive optics. Our staff has over 15 years experience with aspheric CGH nulls. Our goal is to become your single source of responsibility for design and fabrication of diffractive null optics.

How They Work

Our CGH Cylinder Nulls are cylindrical zone plates with a linear carrier perpendicular to the cylinder axis. They produce a cylindrical test beam from the collimated test beam of your Fizeau or Twyman-Green interferometer.

Each CGH Cylinder Null is surrounded by an annular aperture linear reflective grating which diffracts the collimated test beam back onto itself. By turning adjustment screws to remove the annular tilt fringes, the CGH null is optically aligned to your interferometer test beam.

CGH Cylinder Nulls are switched between positive and negative focal length by reversing the tilt of the collimated input beam. Both diffracted orders exist and the correct wavefront must be selected.

Easy to Use

Step 1:

Mount CGH Cylinder Null in collimated test beam with logo facing test optic.

Step 2:

Tilt 1 degree c.c.w. about y-axis for converging test beam or 1 degree c.w. for diverging test beam. Adjust tilt-tip to null annular alignment fringes.

Step 3:

Adjust test cylinder tilt, tip, focus and clock to acquire and null interferogram.

PATENT PENDING

Accuracy

Test accuracy is limited by CGH pattern errors, substrate error (T.W.D.), and interferometer errors. Wavefront distortion of our CGH substrates is measured before patterning. After patterning and phase etching, this measurement requires a different wavelength.

System errors of your interferometer and transmission flat may be compensated in the usual manner—by testing a known flat and subtracting the result from subsequent interferograms.

Interferogram Visibility

Our CGH Cylinder Nulls are phase gratings with a nominal diffraction efficiency of 40 percent—16 percent in double pass. Interferogram visibility is suitable for test optic reflectivities ranging from 1 to 100 percent, peaking for a reflectivity of about 20 percent.

For many 4-inch interferometers, brightness will be reduced when zoomed to an aperture smaller than about Ø30 mm.

Six Models Available

We offer six standard CGH Cylinder Nulls for use with 632.8 nm wavelength interferometers. Model H45F1.5C is e-beam mastered on commercial photo-mask blanks. Other models are contact printed from e-beam submasters onto precision substrates.

Mechanical interference with the CGH frame will reduce the maximum convex cylinder radius from B.F.D. to Rmax for test cylinders larger than the CGH aperture.

Meeting Your Needs

We can provide custom CGH Cylinder Nulls for other F/#'s or wavelengths. Please allow us the opportunity to address your aspheric testing requirements.

Model No.	Aperture	F/#	B.F.D.	Rmax	Frame	Pattern	T.W.D.
H45F1.5C	Ø45 mm	F/1.5	165 mm	64 mm	50R	0.18 wv	0.10 wv
H45F2.4C	Ø45 mm	F/2.4	1105 mm	104 mm	50R	0.12 wv	0.1 wv
H45F4.5C	Ø45 mm	F/4.5	1201 mm	198 mm	50R	0.08 wv	0.1 wv
H45F8C	Ø45 mm	F/8	1360 mm	357 mm	50R	0.06 wv	0.1 wv
H95F3C	Ø95 mm	F/3	1281 mm	279 mm	120	0.12 wv	0.2 wv
H95F6C	Ø95 mm	F/6	1568 mm	566 mm	120	0.07 wv	0.2 wv

Diffraction International

CGH & BINARY OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

CGH
Photomasks

At Diffraction International, we specialize in diffractive optics. Our staff has over 15 years experience in CGH and binary optics technology. Our goal is to become a leading supplier of custom precision diffractive optics. As a small business, we offer flexibility, attention and fast response.

General Description

CGH Photomasks are the basis of our diffractive optic products and are also used directly as diffractive optical elements. CGH Photomasks are encoded using our HoloMask software and written using a MEBES e-beam lithography system. The MEBES is the standard photomask tool in the microelectronics industry. In resolution, accuracy and format size, it is unsurpassed for CGH writing. Unfortunately, the MEBES pattern vocabulary is limited to trapezoids. Our HoloMask software accurately encodes a diffractive optic element using a minimum number of these trapezoids.

Resolution and Feature Size

Patterns are written in a raster manner. The range of address resolution, or pixel size, is 0.10 to 1.1 micron. Common address sizes are 0.50, 0.25 and 0.10 micron.

ON-AXIS

OFF-AXIS

GENERAL

$$\phi(x,y) = \frac{C^2}{1 + \sqrt{1 - (K+1)C^2}} + \sum_i A_i r^i + \sum_j P_j x^j y^j + \sum_l Z_l U_l^i(r, \theta) + \text{custom}$$

Because of exposure proximity effects, the minimum feature size should be not less than 5 pixels. This is seldom a limitation for CGH patterns since a minimum of 10 pixels per grating period (5 pixels per line or space) is required to achieve a typical recording accuracy of 0.1 fringe. Feature sizes below 0.7 microns are extremely difficult to control in processing.

Aperture

Within a rectangular data field, apertures of any shape may be encoded. Patterns may not exceed the MEBES physical reticle limit of 155x155 mm.

Optical Prescription

The optical phase function is specified by an equation. Virtually any functional form $\phi(x,y)$ can be accommodated. Simple zone plate lenses may be specified by giving the focal length and wavelength. More general phase functions are described by radial or cartesian monomials or by Zernike polynomials. We will accept Sweatt model wavefront prescriptions. We can also write a custom wavefront evaluation subroutine or link to your code.

Most often, phase function coefficients are generated by an optical design program. We are familiar with Code V,

Oslo, ZEMAX and GENII formats. If your phase function requires a large number of coefficients, we ask that you supply it on floppy disk.

Units and Dimensions

We usually design in millimeters but can accommodate any unit of measure. Wavelength is in microns or nanometers. The MEBES works in microns.

Artwork Generation

In addition to the phase function, other image elements such as fiducials, title blocks, etc. may be included. A conversion program for customer generated AutoCAD DXF files is available, or we can produce CAD layouts from your sketches. Points and zero width lines are ignored. Text is converted to a monospace font.

Pattern Composition

Multiple images may be combined to form an overall pattern. Images may overlap and can be either positive (digitized area exposed/clear) or negative (digitized area unexposed/dark). It is difficult to add dark features within a positive pattern or clear features within a negative pattern. The MEBES writes most efficiently when patterns repeat in a rectangular array. This is often true of lenslet arrays or linear gratings.

CGH Photomasks

LENSLET ARRAY

Pattern Accuracy

Overall pattern accuracy is determined by encoding, digitization, and registration errors. The registration accuracy of the MEBES is better than 0.15 microns (3 σ) for an address size of 0.25 micron. Allow for somewhat less accuracy when using larger address sizes. Encoding accuracy is user specified in phase and/or linear units. Tighter accuracy specs will result in larger data sets and longer write times.

Complexity

Because of the inefficiency of the MEBES data format in describing CGHs, greater time may be required to input and rasterize the data than in actual writing. It is relatively easy to specify a CGH pattern which, while possible, would exceed most budgets. Pattern complexity depends on a number of factors including pixel size, encoding accuracy, and encoding algorithm. For off-axis CGHs, encoding efficiencies result when fringes are oriented parallel to the y axis. We strive to select the most cost effective encoding parameters to meet your requirements.

Verification

Our encoding and digitization algorithms are rigorously verified through both internal and external checks. Violations of user specified encoding tolerances are reported to a log file.

Phase function coefficients are echoed to a log file. A table of phase function values at customer specified aperture locations can also be generated.

Test plots of any pattern region can be provided. Please specify plot scale. We offer AutoCAD DXF plots on floppy disk, letter size 300 dpi laser plots, or Versatec plots to 40 inch on vellum.

Nikon 2i marks are usually included in CGH patterns. Their locations may be measured to verify positional metrology.

Substrates

We use synthetic fused silica photomask blanks because of their ultra-low thermal expansion, high index homogeneity, and superior etching properties. These are coated to an optical density of 3.0 (950 Å) with either anti-reflective (11%) or bright (55%) chrome.

Photomask sizes are listed below. We use commercial blanks of the highest available flatness grades. The flatness spec refers to the face side, although the back side is of similar quality. Flatness is better over smaller apertures.

For critical applications, we offer custom polished 5009 substrates in an

ultra-precision flatness grade (Flatness Code 01) with a transmitted wavefront distortion of $\lambda/10$ P-V at 633 nm. These serialized substrates are supplied with transmission interferograms.

Coatings

Dielectric anti-reflective coatings can be applied to the back side of CGH photomasks. A coating is standard with our custom polished 5009 substrates.

Cutting, Edging and Mounting

CGH photomasks can be cut, centered and edged to customer specifications. We offer standard or custom mounting.

Prices and Delivery

Please call or fax us for a quotation. Single quantity CGH photomask prices begin at \$1950. Replicas are relatively less expensive. Standard delivery is 3 weeks. Rush delivery is often available at a modest premium.

Additional Literature

- Binary Optics
- CGH Frames and Mounts
- CGH Null Adapter

"Desktop computer encoding of electron-beam written holograms," in *Computer-Generated Holography II*, SPIE Vol. 884, 23-26 (1988), S. M. Arnold.

"E-beam fabrication of computer generated holograms," *Optical Engineering* 24(5), 803-807 (1985), S. M. Arnold.

Size Code	Nominal Size (inch)	Size (Inch)		Size (mm)		Thickness (Inch)		Thickness (mm)		Maximum Aperture		
		Min	Max	Min	Max	Min	Max	Min	Max	Area (mm)	Flatness	Code(s)
4006	4.0 x 4.0 x 0.06	3.97	4.00	100.84	101.60	0.056	0.064	1.40	1.60	Ø75	3.5 μ m	6C
4009	4.0 x 4.0 x 0.09	3.97	4.00	100.84	101.60	0.086	0.094	2.20	2.40	Ø75	2 μ m	2C
5009	5.0 x 5.0 x 0.09	4.97	5.00	126.24	127.00	0.086	0.094	2.20	2.40	Ø100	1 μ m	1C
6009	6.0 x 6.0 x 0.09	5.97	6.00	151.64	152.40	0.086	0.094	2.20	2.40	Ø125	2 μ m	2C
6025	6.0 x 6.0 x 0.25	5.97	6.00	151.64	152.40	0.246	0.254	6.25	6.45	Ø125	1 μ m	1C
7025	7.0 x 7.0 x 0.25	6.97	7.00	177.04	177.80	0.246	0.254	6.25	6.45	Ø150	2 μ m	2C

Diffraction International

CGH & BINARY OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

Binary Optics

At Diffraction International, we specialize in diffractive optics. Our staff has over 15 years experience in CGH and binary optics technology. Our goal is to become a leading supplier of custom precision diffractive optics. As a small business, we offer flexibility, attention and fast response.

Binary Optics

Binary CGHs are surface relief diffractive optics with the pattern etched to a nominal phase depth of π . This increases the intensity of all but the zeroth diffracted order by a factor of 4.

The term binary optic is commonly applied to any quantized-phase grating structure produced by multimask lithographic techniques. Additional masks further increase diffraction efficiency. An M mask process yields $N = 2^M$ phase levels.

Fabrication

Fabrication begins with the encoding and e-beam writing of one or more CGH photomasks as described in a separate bulletin. A CGH photomask may become the binary optic, or it may be contact printed to another substrate.

Both ion milling and reactive ion etching (RIE) are used to create surface relief patterns. RIE more readily produces deep, square-bottomed features and uses the chrome pattern as an etch mask. Ion milling yields better etch depth uniformity and lower scatter, but produces less square sidewalls and usually requires a photoresist etch mask. Often, the deciding factor is feature size.

Three methods we use to add a second pattern level are e-beam direct write, laser direct write, and contact photolithography. Quartz photomask substrates are

used for direct writing and the process is equivalent to phase shift mask fabrication for the microelectronics industry.

Compared to VLSI electronics, fabrication of binary optics requires very deep anisotropic etching and stringent alignment tolerances, but localized defects are of much less concern.

Diffraction Efficiency

The scalar theory diffraction efficiency of an N-phase binary grating is

$$\eta = \left[\frac{\sin(\pi/N)}{(\pi/N)} \right]^2$$

More rigorous vector diffraction theory reduces the attainable efficiency for fine period, binary gratings on substrates of low refractive index.

2ND MASK LITHOGRAPHY

CHROME & 90° ION ETCH

STRIP & CHROME ETCH

1ST MASK LITHOGRAPHY

CHROME & 180° ION ETCH

STRIP & CHROME ETCH

4-PHASE CGH FABRICATION

Mask Levels	Phase Levels	Scalar Theory Efficiency
4	1	10.1%
1	2	40.5%
2	4	81.1%
3	8	95.0%
4	16	98.7%

Binary Optics

BINARY 4-PHASE GRATING

For multi-phase CGHs, alignment of subsequent mask levels is the most critical parameter limiting diffraction efficiency. If not aligned to better than 0.08 grating period, a second mask level does more harm than good. Alignment to 0.05 grating period is usually required to justify a second mask level.

Diffraction International uses a software model to predict binary and multi-phase CGH diffraction efficiencies for various values of etch depth, mask misalignment, and linewidth variation.

Dammann Gratings

Dammann gratings are a type of binary phase grating used to produce a one- or two-dimensional array of equal-intensity light spots. Applications include laser beam multiplexing, optical computing, and structured illumination.

Dammann gratings are useful for array sizes up to about 41×41 spots. Above that size, problems with the computation and with the resolution become dominant. There are 2^N possible solutions for an N-th order grating ($2N+1$ spots) and the resolution requirements increase as $N^{3/2}$. Also, small errors in etch depth will cause the central spot to become much brighter than the others.

9 X 9 DAMMANN GRATING

Coatings

Dielectric anti-reflective coatings can be applied to one or both faces of a binary optic. Although coating performance is degraded on the etched face, the coating can still be beneficial if feature sizes are large compared to the coating thickness.

Binary optics may be aluminized to create diffractive reflection optics. The required etch depths on SiO_2 are much less than for transmissive optics. Reflective binary elements can also be created by omitting the ion etch and aluminizing over an etched chrome pattern.

Replication

We offer binary and multi-phase diffractive optics produced by contact photolithography from e-beam masters. Achievable specifications depend on the substrate. Best results are obtained with photomask or wafer-like substrates for which good contact can be achieved between photomask and substrate.

Binary optics can also be replicated by inexpensive compression molding. Masters for replication are made of nickel electroformed on a quartz original. We can direct you to vendors who will replicate our binary masters. While diffraction efficiency is usually maintained in the replication process, wavefront quality is typically somewhat degraded.

Achievable Specifications

We currently offer 2- and 4-phase binary CGHs produced by e-beam direct writing. The 2-phase binary CGHs are available on a variety of photomask substrates including our custom-polished, ultra-flat 5009 plates. The 4-phase CGHs are currently offered only on commercial 6025 quartz photomask blanks. Minimum feature size is 0.5 micron. The overlay tolerance for 4-phase patterns is 0.2 micron. Etch depth uniformity is $\pm 4\%$ over 60 mm.

Diffraction efficiencies of 35% are readily achieved for 2-phase binary optics. Efficiency is quite insensitive to etch depth, although a $\pm 5\%$ variation will cause a $\lambda/20$ P-V wavefront distortion.

We continually seek to expand our binary CGH capabilities in response to customer needs. Please allow us the opportunity to address your requirements.

Additional Literature

- CGH Photomasks
- CGH Frames and Mounts

Zygo Corporation
Middlefield, CT 203/347-8506
Ad Pages 159, Cover IV Handbook

MANUFACTURE CUSTOM ITEMS

Gradient Lens Corp.
Rochester, NY 716/235-2620
Ad Page 448

Ramar Corp., Westford, MA
Zygo Corporation
Middlefield, CT 203/347-8506
Ad Pages 159, Cover IV Handbook

DESIGN/PROTOTYPE

Photonic Integration Research, Inc.
Columbus, OH 614/876-5655
Ad Page 227
Truax Associates, Southington, CT
UltraOptec Inc.
Boucherville, P.Q., Canada
Tel: 514/449-2096
Ad Page 331

335001

INTERFEROMETERS MICHELSON

MANUFACTURE STOCK ITEMS

C.S.O., Grenoble, France
Direct Optical Research Company
Phoenix, AZ 602/955-2826
Ad Page 335

Mark-Tech Laser, Inc., San Jose, CA
Midac Corp., Irvine, CA
Ometron Ltd., London, England
Optra, Inc., Topsfield, MA
Research-Cottrell, KVB/Analect, Irvine, CA

Zygo Corporation
Middlefield, CT 203/347-8506
Ad Pages 159, Cover IV Handbook

MANUFACTURE CUSTOM ITEMS

Direct Optical Research Company
Phoenix, AZ 602/955-2826
Ad Page 335

Gradient Lens Corp.
Rochester, NY 716/235-2620
Ad Page 448

Graseby Specac Ltd., Orpington, Kent, England
Mark-Tech Laser, Inc., San Jose, CA
Ontar Corp., North Andover, MA
Sciencetech, Inc., London, On., Canada
Spectra Optics, Inc., Block Engineering Group,
Marlborough, MA

Zygo Corporation
Middlefield, CT 203/347-8506
Ad Pages 159, Cover IV Handbook

DISTRIBUTOR/SUPPLIER

Odhner Holographics, Orlando, FL

DESIGN/PROTOTYPE

Axiom Analytical Inc., Irvine, CA
Direct Optical Research Company
Phoenix, AZ 602/955-2826
Ad Page 335
PIKE Technologies, Inc., Madison, WI
Truax Associates, Southington, CT

335501

INTERFEROMETERS OTHER THAN THOSE LISTED

MANUFACTURE STOCK ITEMS

Apollo Optics, Inc., Santa Monica, CA
Chapman Instruments, Inc., Rochester, NY
Direct Optical Research Company
Phoenix, AZ 602/955-2826
Ad Page 335

Graton Technology, Inc., GTI/SpectroVision Div.,
Concord, MA
Norland Products Inc.
North Brunswick, NJ 908/545-7828
Ad Page 93

Test Deep General Aspherics

---with Your
Existing
Interferometer

- Accurate
- Affordable
- Easy to Align
- Custom Nulls
in 4-6 weeks

Diffraction International

DIFFRACTIVE OPTICS ■ ASPHERIC METROLOGY ■ DESIGN ENGINEERING

At Diffraction International, we specialize in precision diffractive optics. Our CGH Null Adapters are based on 15 years experience in aspheric metrology.

Please write, phone or FAX us for further information. We welcome the opportunity to discuss your application specifics.

11345 Hwy 7, #421, Minneapolis, MN 55305, Phone or Fax 612-945-9912

INTERNAL DISTRIBUTION

1. D.E. Beck
2. B.E. Bernacki
3. B. B. Bovee
4. J.F. Cook
5. L.B. Dunlap
6. B.M. Evans
7. L.J. Hunt
- 8-14. L. C. Maxey (7)
15. D. W. McDonald
16. W.T. McKeethan
17. H.M. Meyer
18. A.C. Miller
19. G. N. Miller
20. W. P. Painter
- 21-27. J.E. Rogers (7)
28. T.M. Rosseel
29. M. L. Simpson
30. R.A. Smith
31. J. O. Stiegler
- 32-33. Laboratory Records (2)
34. Laboratory Records - RC
35. ORNL Patent Section
36. I&C Division Publications Office
37. DOE Office of Patent Counsel

EXTERNAL DISTRIBUTION

38. S. M. Arnold, Diffraction International, 11345 Highway 7, #421,
Minneapolis, Minnesota 55305-5300
- 39-40. Office of Scientific and Technical Information, PO Box 62, Oak Ridge, TN, 37831