LA-SUB--96-93

SUMNER ASSOCIATES

100 Cienega, Suite D
Santa Fe, New Mexico 87501
Telephone: 505 984-8041 FAX 505 984-3251

RECEIVED

AUG 01 1996

OSTI

FINAL REPORT

0412J0004-3G

TASK 2

May 22, 1996

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

MASTER

DISCLAIMER

**Portions of this document may be illegible
in electronic image products. Images are
produced from the best available original
document.**

Final Report
Sumner Associates
Task 2

April 20, 1996

The task as defined as Task Order 2 pursuant to Subcontract No. 0412J0004-9G between the University of California and Sumner Associates.

The Subcontractor was requested to perform the work described below:

- 1.0 Assist in completion of ongoing projects in ultra-sensitive detection of single molecules, single molecule spectroscopy and electrophoretic analysis
- 2.0 Assist in the preparation of technical reports, manuscripts for publication and conference presentations in areas prescribed in Task 1.0
- 3.0 Provide consultation, as required in the area of ultra-sensitive detection
- 4.0 Attend and assist in presentations, as required, of ultra-sensitive detection results at national and international meetings
- 5.0 Provide written summary of results and status of tasks

The Subcontractor's principal investigator for the subtask order was Dr. E. Brooks Shera.

The work requested in items 1 and 2 above has been completed and is discussed in the attached report "Single Molecule Electrophoresis" by A. Castro and E. B. Shera which has been published. See Attachment 1.

The work requested in items 3 and 4 above has involved several facets which have been the subject of internal Los Alamos National Laboratory reports. (Copies available upon request.) A major component of this work is reported in the document "Single Molecule Detection: Applications to Ultrasensitive Biochemical Analysis" by A. Castro and E. B. Shera. See attachment 2. A further component of this work has involved discussions and research leading to the development of a CRADA between the Laboratory and Pioneer Hi-bred International, a major agricultural supplier. A copy of this CRADA is available upon request.

Numerous other written reports have been provided during the course of this work in compliance with item 5. In the interest of brevity copies of the most succinctly edited reports are attached.

Attachment 1

Single-Molecule Electrophoresis

Alonso Castro* and E. Brooks Shera

Biophysics Group, Mail Stop D434

Los Alamos National Laboratory

Los Alamos, NM 87545

ABSTRACT

A novel method for the detection and identification of single molecules in solution has been devised, computer-simulated, and experimentally achieved. The technique involves the determination of electrophoretic velocities by measuring the time required by individual molecules to travel a fixed distance between two laser beams. Computer simulations of the process were performed beforehand in order to estimate the experimental feasibility of the method, and to determine the optimum values for the various experimental parameters. Examples of the use of the technique for the ultrasensitive detection and identification of rhodamine-6G, a mixture of DNA restriction fragments, and a mixture of proteins in aqueous solution are presented.

* To whom correspondence should be addressed.

INTRODUCTION

The rapid and efficient detection and separation of minute quantities of biologically important molecules plays a central role in a variety of fields, including molecular biology, biotechnology, immunology, medical diagnosis, and forensic analysis. It has proven difficult to identify and separate biomolecules at such low concentrations by existing means. Therefore, it is important to develop methods that are capable of probing samples at such low concentrations with adequate sensitivity, resolution, quantitation, and ease. Here, we describe a new method for the ultrasensitive detection and identification of fluorescent compounds in solution. The technique involves the measurement of electrophoretic velocities of individual molecules in a mixture, and identification by comparison with the electrophoretic velocity known to be characteristic of a particular molecular species. In our experiments, the sample solution is contained in a capillary cell at a concentration in the femtomolar range. An external voltage is applied between the ends of the capillary, which causes the molecules to migrate towards the cathode or anode, depending on their charge. The migration velocity of each molecule, which is the vector sum of the electrophoretic velocity and the electroosmotic flow velocity, is calculated from the time required to travel a fixed distance between two laser beams. The passage of individual molecules through each of the laser beams was detected by using a modified version of our recently developed technique of single fluorescent molecule detection.¹⁻³ A Monte Carlo computer simulation of the electrophoretic separation and single-molecule detection processes was performed in order to demonstrate the feasibility of the method, and to find the optimum parameters for the experimental execution. The experimental application of the method to the ultrasensitive detection and identification of a mixture of TOTO-stained lambda DNA restriction fragments, a mixture of phycobiliproteins, and the single-fluorophore rhodamine-6G has been demonstrated

in this work. Although we have focused on the detection and identification of biologically important molecules, we believe that the technique will also find applications in many areas of ultrasensitive organic and inorganic chemical analysis.

EXPERIMENTAL SECTION

Single-Molecule Electrophoresis (SME). The basic experimental arrangement is depicted in Fig. 1. A Spectra-Physics 3800 frequency-doubled mode-locked Nd:YAG laser producing 70-ps pulses at 532-nm wavelength and 82 MHz repetition rate was used as the excitation source. The laser output was attenuated by a variable neutral density filter, and then split into two parallel beams by a beamsplitter-mirror combination. The resulting 1-5 mW laser beams were focused into the sample capillary cell by an achromatic lens to yield two 10- μm spots ($1/e^2$ value). The ends of the 2-cm long, 100- μm square cross-section glass capillary were connected to two sample reservoirs by means of plastic tubing (Fig. 2). This arrangement, combined with a 10 x 10 μm detection area (see below) translates to a 1% sampling of the total number of molecules flowing through the capillary. The reservoirs, capillary, and tubing were filled with the sample solution. The total sample volume required to fill the sample compartment is about 3 mL. For the purpose of accommodating samples present in minuscule amounts, the sample compartment of our apparatus could be reconstructed, if necessary, to have a volume as small as 1 μL . The output of a high voltage power supply was connected to the sample reservoirs by means of platinum electrodes. Typical applied voltages were on the 100-200 V/cm range. The applied electric field causes the ultra-dilute solute molecules to migrate to the cathode or anode, depending on their charge. As the individual molecules move through each laser beam due to the electrophoretic effect, repeated excitation-emission cycles produce a fluorescence photon burst. This fluorescence light from each beam is

collected by a 40×0.75 N.A. microscope objective and spatially filtered by a 0.4×0.4 mm square slit, which define a $10 \times 10 \mu\text{m}$ detection area per laser beam. The light is then spectrally filtered by 30 nm-bandwidth eight-cavity interference filters centered at the wavelength of emission maximum, and detected by two EG&G single-photon avalanche photodiodes. Each detector output signal is analyzed by independent time-correlated single-photon counting electronics under computer control. The detection electronics reject Raman and Rayleigh scattering by using a time-gate window set such that only delayed fluorescence photons are detected, thus increasing the signal-to-noise ratio of single-molecule detection.¹ Every burst produced at the second beam by a given molecule will be delayed with respect to the burst produced at the first beam by a time equal to the inter-beam distance divided by the molecule's migration velocity. Therefore, a histogram of these times will produce a series of peaks, one for each molecular species present, since different species generally exhibit different electrophoretic mobilities. A sample run time of 819.2 seconds was used in the present measurements.

Sample preparation. Lambda DNA (full-length, *Apa I* and *Xba I* digests) and TOTO-1 were purchased from New England Biolabs (Massachusetts), and Molecular Probes (Oregon), respectively. Nucleic acid solutions were prepared by dilution in TBE buffer (10 mM tris, 9 mM boric acid, 0.1 mM EDTA, pH 8.0) to the desired concentration. TOTO-1 was added to the DNA samples at a final base pair to dye ratio of 5:1. B-phycoerythrin and B-phycoerythrin-streptavidin-conjugate were purchased from Molecular Probes. Protein solutions were prepared by dilution in 10 mM potassium-hydrogen-phthalate buffer solution, pH 4.8. Rhodamine-6G was purchased from Exciton (Ohio) and diluted in water. Water was deionized and doubly distilled prior to use. All solutions were prepared immediately before each experiment.

COMPUTER SIMULATIONS

A Monte Carlo modeling computer program for the simulation of the single-molecule electrophoresis process was developed on the basis of the above description. This computer simulation has proven to be an invaluable tool for predicting the feasibility of the single-molecule electrophoresis experiments, as well as for understanding the various factors affecting the molecular migration and low-level fluorescence detection processes. The algorithm consists of generation of individual molecules at random positions and times at the top of the capillary cell, translation of molecules due to both electrophoretic-electroosmotic migration and random diffusion, excitation of molecules when they interact with the gaussian laser beams, fluorescence emission, photon collection and detection, and random generation of background noise. The simulation of the electrophoretic migration consisted of adding a measured constant velocity to each generated molecule in the direction of the capillary axis. The simulation of the single-molecule detection process employed the common fluorophore rhodamine-6G as model system, and considered such parameters as laser power, extinction coefficient, photodegradation quantum efficiency,⁴ fluorescence quantum yield, and overall photon detection efficiency. Simulations were carried out in a Cray Y-MP computer. Results were output as the cross-correlation between the raw signals from the two detection channels, such that a peak is obtained for each molecular species present in the sample. Alternatively, the results were output as a histogram of travel times, where, again, each species is represented by a peak. In this case, the time position of each fluorescence burst that exceeded a set threshold was determined by least-squares fitting to a gaussian curve. The travel time for a given molecule was then calculated by subtracting the time positions of the fluorescence bursts for that molecule originating at each of the two laser beams. This latter data analysis method allows for greater histogram peak

resolution at the expense of decreased histogram peak amplitudes. Some of the simulation parameters, such as laser power, electrophoretic velocity (proportional to the applied voltage), and distance between laser beams were varied until the best signal-to-noise ratio was obtained, and were later used experimentally.

In order to investigate the effect of molecular diffusion on the efficiency of electrophoretic velocity determination, a set of simulations was performed where the aqueous-solution diffusion coefficient was varied from that corresponding to a small common fluorophore, such as rhodamine-6G ($\sim 3 \times 10^{-6} \text{ cm}^2/\text{sec}$),⁵ to that of a large biomolecule, such as a 6.6 Kbp nucleic acid ($\sim 3 \times 10^{-8} \text{ cm}^2/\text{sec}$).⁶ Figure 3 shows the results. As expected, the cross-correlation width increases as the diffusion coefficient value is increased, because diffusion causes a broadening in the distribution of total molecular velocities. A decrease in peak amplitude (and peak area) is also observed because diffusion causes a deviation of the molecular trajectories from a straight path between the two laser beams.

The dynamic range of the technique was investigated by performing another set of computer simulations. The sample concentration was varied over four orders of magnitude, from 62.5 attomolar to 1.02 picomolar. The net peak amplitude (and area) was found to increase linearly with concentration over the whole range ($r = 0.9999$). The signal-to-noise ratio increased by only 1300 times over the same concentration range, indicating that the baseline noise increased somewhat with concentration. At high concentrations, the probability of two or more molecules residing between laser beams increases. Thus, there is an increase in the number of "false" migration times originating from two different molecules present simultaneously between the laser beams. If necessary, this slight decrease in signal-to-noise ratio at high concentrations could be overcome by adding a third laser beam and detector, which would allow the use of algorithms that would effectively discard such "false" migration times.

RESULTS

Having succeeded in predicting the feasibility of the single-molecule electrophoresis technique, the experimental demonstration was then accomplished. It is important to note that the computer simulations were performed employing molecular parameters corresponding to a single fluorophore, which represents the ultimate level of detection difficulty. Therefore, the detection of biological molecules, which often contain multiple fluorophores, is expected to proceed with even greater efficiency.

In one series of experiments, a mixture of DNA restriction fragments was chosen for the experimental demonstration of the technique. Since the fluorescence quantum yield of DNA is very small, we intercalated the fluorescent dye TOTO-1 (a dimer of thiazole orange) into DNA.^{3,7} Figure 4 shows the fluorescence bursts observed from the two detection channels for a 10 femtomolar solution of lambda DNA (48,502 base pairs) in TBE buffer to which TOTO-1 was added at a base pair to dye ratio of 5:1. Every burst on the first detector appears shifted on the second detector by the time taken by a individual molecule to travel the distance between the two laser beams. To investigate the identification of DNA fragments according to size, we used a mixture of DNA fragments of three different sizes: 38, 24 and 10 Kbp. It is a well known fact that DNA fragments of different sizes experience essentially the same drag-force per unit charge in free solution, which makes their separation impossible.⁸ However, capillary electrophoresis separations of DNA fragments have been successfully demonstrated by using a semi-dilute, low viscosity, entangled polymer solution as sieving media.^{9,10} We therefore added hydroxy-propyl-methyl-cellulose at a concentration of 0.25% to that effect. The time required for each molecule in the mixture to travel the 250- μ m distance between laser beams was recorded. As expected, larger fragments are retarded by the sieving media, whereas smaller fragments travel faster. Figure 5

shows a histogram of these times for approximately 1000 DNA molecules. Three peaks are observed, corresponding to each one of the DNA fragment sizes. The peak assignment was verified by running each DNA digest independently under identical experimental conditions.

In another series of experiments, a mixture of two fluorescent proteins was analyzed. Phycoerythrin (PE), and streptavidin-conjugated phycoerythrin (SPE) were mixed at an individual concentration of 20 fM in KH-phthalate buffer. These proteins were chosen because they fluoresce in the visible region, and because of the importance of phycobiliprotein conjugates in flow cytometry and immunofluorescence.¹¹ Separation of the two components was enhanced by taking advantage of the difference in their isoelectric points, which were determined by standard agarose-gel isoelectric focusing techniques. The isoelectric point values obtained were 4.43 ± 0.03 for PE,¹² and 4.57 ± 0.03 for SPE. The pH of the buffer solution was adjusted to 4.8, where PE acquires a more negative character compared to SPE. Under these experimental conditions, both proteins will migrate towards the cathode, but SPE will migrate at a lower rate, both because of a less negative character, and because of its larger size. Figure 6 shows a histogram of travel times for approximately 2500 protein molecules, where a peak is observed for each component. Again, peaks identities were determined by running each component independently.

Our computer simulations indicated that our technique is sensitive enough to detect and analyze a small, single-fluorophore molecule, as well as fluorescently-labeled DNA and the phycoerythrins, which are multifluorophore-bearing macromolecules. To verify this hypothesis, we analyzed a sample consisting of 20 fM rhodamine-6G in water. Figure 7 shows the single-molecule fluorescence bursts observed at the two detection channels. Fluorescence bursts corresponding to the same molecule are labeled by the same number. Bursts at detector 1 are delayed by a constant amount of time with respect to the bursts at detector 2 (about 420 msec in the

case illustrated). The molecule in detector 2 marked by an "X" either drifted away so that it missed the detector 1 laser beam, or photodegraded while in the first beam. Water blanks for each detector are also shown, where single-molecule fluorescence bursts are not present. A histogram of the migration times is shown in Fig. 8. These results show the applicability of the technique to the analysis of single-fluorophore molecules.

DISCUSSION

One of the most commonly used techniques for DNA restriction fragment sizing and for protein separations is gel-slab electrophoresis. Despite its popularity, gel electrophoresis suffers from some limitations, mainly because it involves a series of manually intensive procedures that cannot be run unattended and cannot be readily automated; casting gels, applying samples, and running and subsequently staining the gels are time-consuming tasks susceptible to poor quantitative accuracy and poor reproducibility. In some cases, in order to improve sensitivity, a radioisotope needs to be incorporated into the sample, which brings up a set of safety and environmental concerns. An additional advantage over conventional gel electrophoresis is that the identification process is continuous, rather than batch oriented, and the process is therefore suitable for applications that require rapid, continuous on-line analysis. Capillary zone electrophoresis (CZE) alleviates some of these problems, offering rapid, automated analysis, improved reproducibility, and better quantification. Our technique shares with CZE the basic idea of measuring migration times for performing the analysis. In CZE, all molecules are simultaneously injected at one end of the capillary, and the sample components separate into bands according to electrophoretic velocities. The detector at the end of the capillary records the arrival of the various bands. In SME, the migration times are recorded for each molecule

arriving at the laser beams in a continuous way, as opposed to the batch mode of operation of CZE. Thus, SME is also suitable for applications that require continuous, real-time analysis, such as those processes where the sample composition and concentration change with time. SME, however, does not currently achieve the resolution of gel-slab or capillary electrophoresis ($<1\%$). The present resolution limit of the technique can be determined from the widths of the peaks shown in figure 3. For smaller diffusion coefficients, such as those associated with biomolecules, the resolution values (expressed as peak FWHM divided by peak center position) lie in the range 2-3%, which is satisfactory for most laboratory separations. In principle, increasing the inter-beam separation would cause an increase in the technique's resolution. Since the root-mean-square molecular displacement due to diffusion scales as the square-root of time, increasing the total distance will cause a decrease in the electrophoretic velocity distribution width. We are presently in the process of further investigating this matter.

It is conventional to quote detection sensitivity in terms of a "limit of detection" parameter (LOD) equal to the number of molecules required to produce a signal equivalent to three times the standard deviation of the baseline (background) of the measurement. In the present measurements, we detect individual molecules as they pass through the laser beams. Therefore, the LOD is clearly lower than 1 molecule. Still, it is useful to present LOD values as a measure of signal and background levels and for comparison with other published results. For the three solutions discussed in the experimental section, LOD values computed from the raw detector data streams are: nucleic acids - 0.047 molecules, proteins - 0.052 molecules, and rhodamine-6G - 0.28 molecules. This latter value can be compared with the recent result of Chen et al.,¹³ who obtained an LOD of 6 molecules for sulforhodamine-101 using laser-induced fluorescence capillary zone electrophoresis coupled with a sheath flow system.¹⁴

Although the present technique has a superior LOD compared with CZE, a more important property not revealed by the LOD values is the ability to analyze extremely dilute solutions. By its nature, CZE is a batch-oriented technique and is limited by the small volumes that can be loaded onto the capillary. A number of molecules equal to or greater than the LOD must be present in this small sample volume, or the technique fails. SME, on the other hand, is a continuous analysis process in which arbitrarily large volumes of solution can be analyzed if necessary to identify a rare or extremely dilute component, or to obtain statistical accuracy in determining the concentration of such a component. The ultimate dilution that can be analyzed is limited only by the running time available.

In the present experiments, we have not attempted to optimize the efficiency with which the sample solution volume is used. Thus, the capillary tube is much larger than the small laser beam spots and therefore most sample molecules pass by undetected. A smaller capillary, larger beam spots, or a sheath flow focusing system would make more efficient use of the sample. Nevertheless, it is useful to explore the limits of the present SME system for the analysis of dilute solutions. For this purpose, we assume that the data from the detectors are analyzed using the cross correlation technique, which is convenient for long runs and large data sets. For the present experimental conditions, we estimate that with a 1 fM solution run for 35 minutes, approximately 80 molecules will pass through the detection regions. For rhodamine-6G, eighty molecules yield a cross correlation signal amplitude of about 2232, or 27.9/molecule. This signal is linearly proportional to the concentration. The cross correlation background is approximately 41.6 times the runtime in seconds. These values lead to the following relationship:

$$S/N \text{ ratio} = 0.16 \times \sqrt{\text{runtime(sec)}} \times \text{concentration (fM)}.$$

Thus, for a given concentration, the S/N ratio will improve as the square root of the time the sample is run. Figure 9 shows a plot of the runtime required to obtain a S/N = 3 as a function of solution concentration. Note that concentrations below 1 fM can be analyzed in a few minutes and concentrations above 5 fM can be analyzed in only a few seconds. A concentration limit of 1 fM is more than 100 times smaller than state-of-the-art CZE.¹³

Although the single-molecule electrophoresis technique relies on measuring molecular fluorescence, non-fluorescent molecules may be detected by attaching a fluorescent tagging molecule to them. In addition, some of the experimental conditions such as buffer composition, pH, viscosity, inner-surface capillary coating, excitation and emission wavelengths, among others, can be optimized to achieve the best separation of the particular sample components being analyzed. In fact, many of the analytical protocols specially developed for capillary electrophoresis separations are directly applicable to the present technique. For many years, researchers have optimized various capillary electrophoresis methods for the separation of a large variety of chemical species ranging from small organic and inorganic ions, to various kinds of pharmaceutical drugs and natural products.

The new method described here promises to combine the advantages of free-solution capillary electrophoresis (system automation, speed, reproducibility) with the unsurpassed sensitivity of single-molecule detection. The sensitivity and versatility of the method may open the way to develop fluorescence immunoassay, hybridization, and DNA fingerprinting techniques without the need for extensive DNA amplification using the polymerase chain reaction (PCR)^{15,16} or other methods. Although PCR is a highly effective amplification mechanism, the use of many PCR cycles may introduce ambiguities arising from contamination and by mechanisms not yet fully understood.¹⁶⁻¹⁹ Besides the demonstrated ability for the analysis of single-fluorophores, mixtures of nucleic acids and of proteins, the technique may find

applications in many other fields that require the ultrasensitive analysis of sample components.

ACKNOWLEDGEMENT

We thank Lloyd Davis for helpful discussions and Mark Peters for technical assistance.

REFERENCES

1. Shera, E. B.; Seitzinger, N. K.; Davis, L. M.; Keller, R. A.; Soper, S. A. *Chem. Phys. Lett.* **1990**, *174*, 553-557.
2. Soper, S. A.; Davis, L. M.; Shera, E. B. *J. Opt. Soc. Am. B* **1992**, *9*, 1761-1769.
3. Castro, A.; Fairfield, F. R.; Shera, E. B. *Anal. Chem.* **1993**, *65*, 849-852.
4. Soper, S. A.; Nutter, H. L.; Keller, R. A.; Davis, L. M.; Shera, E. B. *Photochem. Photobiol.* **1993**, *57*, 972-977.
5. Tamai, N.; Yamazaki, T.; Yamazaki, I.; Mataga, N. *Chem. Phys. Lett.* **1985**, *120*, 24-28.
6. Madelung, O., Ed. *Landolt-Börnstein - Numerical Data and Functional Relationships in Science and Technology, New Series*, Springer-Verlag: Berlin, 1990; vol 1c, p. 272.
7. Rye, H. S.; Yue, S.; Wemmer, D. E.; Quesada, M. A.; Haugland, R. P.; Mathies, R. A.; Glazer, A. N. *Nucleic Acids Res.* **1992**, *20*, 2803-2812.
8. De Gennes, P. G. *J. Chem. Phys.* **1971**, *55*, 572.
9. Yoshinobu, B.; Sumita, C.; Hide, K.; Ishimaru, N.; Samata, K.; Tanaka, A.; Tsuhako, M. *J. Liq. Chromatogr.* **1993**, *16*, 955-965.
10. Zhu, M.; Hansen, D. L.; Burd, S.; Gannon, F. *J. Chromatogr.* **1989**, *480*, 311-319.
11. Haugland, R. P. *Handbook of Fluorescent Probes and Research Chemicals*, Molecular Probes, Inc.: Eugene, OR, 1992.
12. In agreement with the literature value of 4.38. Glazer, A. N.; Hixson, C. S. *J. Biol. Chem.* **1977**, *252*, 32-42.
13. Chen, D. Y.; Adelhelm, K.; Cheng, X. L.; Dovichi, N. J. *Analyst* **1994**, *119*, 349-352.
14. Pinkel, D.; Stovel, R. In *Flow Cytometry: Instrumentation and Data Analysis*; Van Dilla, M. A., Dean, P. N., Laerum, O. D., Melamed, M. R., Eds.; Academic Press, Inc.: London, 1985; Chapter 3, pp 77-128.

15. Scharf, S. J.; Horn, G. T.; Erlich, H. A. *Science* **1986**, *233*, 1076-1078.
16. Bej, A. K.; Mahbubani, M. H.; Atlas, R. M. *Crit. Rev. Biochem. Biophys.* **1991**, *26*, 301-334.
17. Reiss, J.; Krawczak, M.; Schloesser, M.; Wagner, M.; Cooper, D. *Nucleic Acids Res.* **1990**, *18*, 973-978.
18. Dunning, A. M.; Talmud, P.; Humphries, S. E. *Nucleic Acids Res.* **1988**, *16*, 10393.
19. Kwoh, S.; Higuchi, R. *Nature* **1989**, *339*, 237.

FIGURE CAPTIONS

Figure 1. Schematic diagram of the experimental set-up.

Figure 2. Single-molecule electrophoresis set-up. An expanded view of the square-bore capillary cell is depicted at left, showing the two parallel laser beams focused through the cell. Fluorescence light, represented by two sets of arrows, is collected at right angles.

Figure 3. Cross-correlation traces between the simulated signals from the two detection channels for various diffusion coefficient values. The peaks x-axis center position represents the most probable time taken by a molecule to travel the distance between the laser beams. Diffusion coefficient units are cm^2/s .

Figure 4. Fluorescence signals for a 10 fM lambda DNA solution in TBE buffer. Each fluorescence burst represents the passage of a molecule through the laser beam. The bursts generated at the second laser beam are delayed by a fixed amount of time, about 600 msec in this particular case. Applied voltage: 120 V/cm.

Figure 5. Histogram of travel times for a mixture of three different DNA fragments in TBE/HPMC buffer (10 fM each). Applied voltage: 150 V/cm.

Figure 6. Histogram of travel times for a mixture of 20 fM phycoerythrin and 20 fM streptavidin-phycoerythrin in KH-phthalate buffer. Applied voltage: 100 V/cm.

Figure 7. Fluorescence signal versus time for detector 1 (top) and detector 2 (bottom) for a 20 fM rhodamine-6G aqueous solution. In this case, molecules travel towards the cathode, such that they are first observed in detector 2, and then in detector 1. Fluorescence bursts corresponding to the same molecule are labeled by the same number. The molecule in detector 2 marked by an "X" either drifted away so that it missed the detector 1 laser beam, or

photodegraded while in the first beam. Water blanks for each detector are also shown. Applied voltage: 100 V/cm.

Figure 8. Histogram of travel times extracted from Fig. 7. The distribution peak appears at negative times, because our sign convention considers that positive times correspond to migration towards the anode.

Figure 9. Estimated running time required to achieve a $S/N = 3$ versus concentration for the single-fluorophore rhodamine-6G.

Attachment 2

Reprint removed
KJ