RECEIVED BY DTIC MAR 17 1970

COO 1500-17

RADON PROGENY INHALATION STUDY

AS APPLICABLE TO URANIUM MINING

MASTER

**FIFTH ANNUAL PROGRESS REPORT on
U. S. Atomic Energy Commission Contract No.**

AT (11-1) - 1500

For Calendar Year 1969

COLORADO STATE UNIVERSITY - FT. COLLINS, COLO.

DISCLAIMER

**Portions of this document may be illegible
in electronic image products. Images are
produced from the best available original
document.**

Fifth Annual Progress Report on
U. S. Atomic Energy Commission
Contract No. AT(11-1)-1500

for research entitled

RADON PROGENY INHALATION STUDY

As Applicable to Uranium Mining Operations

covering the period
Calendar Year 1969

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED
NK 2-23-95

Principal Investigator

Keith J. Schiager
Department of Radiology and Radiation Biology
Colorado State University, Fort Collins, Colorado

Submitted February 28, 1970

NOTICE

This report contains information of a preliminary nature and was prepared primarily for internal use at the originating installation. It is subject to revision or correction and therefore does not represent a final report. It is passed to the recipient in confidence and should not be abstracted or further disclosed without the approval of the originating installation or DTI Extension, Oak Ridge.

LEGAL NOTICE

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:
A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.
As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.

~~DISTRIBUTION OF THIS DOCUMENT IS LIMITED~~
~~To Government Agencies and Their Contractors~~

ACKNOWLEDGEMENTS

Data for this report from the Ambrosia Lake area of New Mexico were made possible through the cooperation of the mine operators and personnel in the New Mexico Inspector of Mines Department. Messrs. M. H. Ward, Royce Wyrick, and "Ed" Kanizar of United Nuclear Homestake Partners are especially thanked for arranging locations, transportation, hoistmen, and a helper for seven days of experiments in the Homestake Section 25 mine.

Messrs. Howard H. Robinson, Joe D. Longacre Jr., and William H. Hayes, acting State Inspector of Mines, were very cooperative in allowing use of a small diesel generator underground.

TABLE OF CONTENTS

Section	Page
Acknowledgements	i
Table of Contents	ii
I. Introduction	1
II. Background of Overall Project	1
III. Development of Personnel Background	3
IV. Preparation of Equipment for Taking Data in Uranium Mines	4
V. Research Data from United Nuclear Homestake Partners Section 25 mine	5
VI. Updating of Computer Programs	10
VII. Development of a Size Discriminating Thermal Precipitator	16
VIII. Development of a Direct Reading, Working Level Survey Meter	22
IX. Bibliography	32

Figure Number

1. Counting Equipment in Mine Location	6
2. Diagram of Portion of United Nuclear Homestake Partners Section 25 mine sampled	8
3. Radon Concentrations at Various Sampling Sites	11
4. Measured Working Levels at Various Sampling Sites	12
5. Penetration (F) through Diffusion Tubes	15
6. Sketch of Thermal Precipitator	17
7. Cross Section of Thermal Precipitator	18
8. Contribution of Individual Radon Daughter Products to Working Level Concentrations	24
9. Survey Meter Response (cmp per WL)	27
10. Direct Reading, Working Level Survey Meter	28
11. Count Rate Meter Circuit Diagram for Direct Reading WL Survey Meter	30
12. Count Down Scaler Circuit Diagram for Direct Reading WL Survey Meter	31

Table Numbers	Page
I. Radon Progeny Concentration Summery	13
II. Calculated Response of Survey Meter	26

I. INTRODUCTION

The main efforts of the past year have been directed towards investigation of the behavior of the radioactive aerosols in actual mining operations. Activities can be categorized as follows:

1. Preparation of equipment for taking data in uranium mines.
2. Development of background of personnel to carry out intended experiments.
3. Research data from United Nuclear Homestake Partners Section 25 mine in the Ambrosia Lake area near Grants, New Mexico.
4. Updating of computer programs for data analysis in order to extract more information.
5. The development of a thermal precipitator for particulate size analysis in the submicron range.

II. BACKGROUND OF OVERALL PROJECT

The broad objectives of the project remain the same as originally stated in the First Annual Progress Report of February 1966, namely:

1. To investigate the general relationships between physical characteristics of particulate materials carrying radon daughter products and the deposition and retention of these materials in the human respiratory tract.
2. To investigate the relationships between airborne activity measurements made with various kinds of field survey instruments and the actual exposure hazards from radon and its daughter products.
3. To investigate the effectiveness of various mine atmosphere control measures applicable to uranium mines as well to similar deep mines.

Diffusion tube analyses made by Holleman (1) indicated that the majority of aerosol particles to which radon daughters attach were less than .067 micron in effective aerodynamic diameter. Since little information is available on the relationship of size and lung deposition, a mine chamber was built at Colorado State

University to simulate mining atmospheres. This was necessary not only for convenience but also because of the short half-life of Radium A. In actual mining operations miners cannot get out of the mine, shower, and get to the whole body counter in a short enough time for the equipment to obtain a high enough count rate to be significant for analysis. A second factor is that mines today have better ventilation and as a consequence miners have a lower body burden and hence a lower count rate at all times.

In order to ensure that mine atmospheres are actually simulated in the chamber, one must have actual mine data. The main thrust of the project now is therefore to determine the characteristics (principally concentrations and aerosol size distributions) of the radon progeny as functions of:

1. The type of uranium mine (ore grade and rock type).
2. The concentrations of other airborne materials (diesel smoke, water vapor and condensation nuclei).
3. The ventilation control methods (flow rates, distance from fresh air supply, filtration and/or recirculation, drift cross sectional area, etc.).

For particles in the sub-micron aerodynamic diameter range, diffusion tubes have been the principal means of classification of diameters, based on theory developed by Gormley and Kennedy (2). For a monodisperse aerosol the diffusion coefficient as measured by diffusion tubes is independent of flow rate. Pollak and Metnieks (3) have shown that the measured diffusion coefficient is a function of flow-rate for polydisperse aerosols, and further that the diffusion coefficient measured at very low flow rates is associated with the largest aerodynamic diameter of the aerosol population. Fuchs et al (4) give a method for determining the size distribution of polydisperse aerosols by the diffusion method assuming log normal distributions. We are using diffusion tubes for particle size analysis and are writing a computer program for the analysis (see revision of computer program section VI).

We are building an instrument to size submicron aerosols using the principle of thermal precipitation. It has been shown by Fuchs (5) and Walkenhorst (6) that polydisperse aerosols less than 2 microns in aerodynamic diameter can be segregated using a thermal precipitator. We intend to increase this effect by taking advantage of

several parameters which will enhance separation (see the development of a thermal precipitator section VII).

III. DEVELOPMENT OF PERSONNEL BACKGROUND.

Former junior investigators Dan F. Holleman and David E. McCurdy have completed their Ph.D. degrees and are no longer at Colorado State University. The present junior investigators are relatively new to the RADON PROGENY INHALATION STUDY, and have had to devote considerable time to developing background in order to pursue the study intelligently. Ronald L. Beethe and John E. Doerges attended an Aerosol Research Meeting at the Health and Safety Laboratory, U.S. Atomic Energy Commission in New York City on April 29-30, 1969. A follow-up meeting in Albuquerque, New Mexico was also attended by Mr. Doerges during January, 1970.

Three specific areas of study are related to the research project, namely:

1. An advanced graduate course in the mechanics of aerosols was taken by Mr. Doerges during the fall quarter of 1969. This course has been helpful in planning experiments and in the development of the thermal precipitator.
2. Courses in computer programming are being completed by the present junior investigators. Analysis of data will be expedited by computer processing. Programs are being written in subroutine blocks and will be kept on tapes for permanent record and updating of data. The computer programs are being documented so that other users will require a minimum amount of time to comprehend and utilize them. More details on the computer program are contained in section VI.
3. The calibration of the thermal precipitator for size discrimination will be done with polydisperse aerosols. The electron microscope will be employed to determine the geometric mean radius and the geometric standard deviation of the particle deposition as a function of the point of deposit, volumetric flow rate, and temperature differential between the hot and cold plates of the precipitator. With this in mind a course on the theory, and

operation of the electron microscope was taken in the summer of 1969 by John Doerges and a second course in sample preparation is presently in progress.

IV. PREPARATION OF EQUIPMENT FOR TAKING DATA IN URANIUM MINES.

In September of 1968 members of this research group visited several uranium mines in the Ambrosia Lake area near Grants, New Mexico. Dan Holleman, assisted by John Doerges, Ed Powers, and David Miller attempted to obtain diffusion tube measurements from mining areas in the Kerr-McGee Section 23 mine. The closest suitable location for the collection of data from the counting trailer was picked and the filter samples were delivered from the sample site to the counting trailer by runners. Even though we had good cooperation of all concerned, samples could not be delivered before the radium A deposited on the filter had decayed to the point that it was no longer useful because of statistics. Since we are using a spectroscopic method of analysis developed by D.E. McCurdy and D.F. Holleman (7), and the analysis depends on a significant radium A count, the effort was futile.

To overcome the situation we prepared to take the equipment into the mine and do our counting in place. A portable diesel powered generator was purchased to supply power and all necessary equipment was packaged to withstand hauling into and around the mines. A rugged table was constructed to support the equipment. The table could easily be knocked down to facilitate transport. Following is a list of equipment prepared and taken into the Homestake Section 25 mine:

1. Portable diesel powered generator.
2. Four-hundred channel pulse-height analyzer with typewriter output.
3. Power supply and silicone surface barrier detectors.
4. Portable field desk with filters, filter papers, stop watches, slide rule, capillary tubes, and other small items.
5. Two gas flow regulators with flex-tube manometers.
6. Three Gelman pressure/vacuum pumps.
7. Sixteen Lucas chambers.

8. One Gardner condensation-nuclei counter.
9. Diffusion battery.
10. Electrical cable.
11. Table.
12. Portable battery powered TLD air samplers.
13. Dwyer pitot tube anemometer.

Figure 1 is a picture of the counting equipment in a mine location.

We plan to visit several uranium mines and take data for one to two weeks at each mine. The project has two mobile trailers, one a 27 foot trailer house (used at the Schwartzwalder Mine) and the second a mobile counting laboratory. Since the mobile laboratory trailer is very heavy and in vivo counting at the mines is not anticipated in the present experiments, the living room of the trailer house was modified into laboratory work space and the trailer was rewired. All equipment for mine experiments can be stowed in the trailer and a pickup used for towing. Three persons can stay in the trailer at the mine site.

V. RESEARCH DATA FROM UNITED NUCLEAR HOMESTAKE PARTNERS SECTION 25 MINE.

Arrangements were made with Mr. M. H. Ward, Mr. Roy McMillan, and Mr. Royce Wyrick the week of August 27, 1969, to take data in either the Section 23 or 25 mine of United Nuclear Homestake Partners. Arrangements were also made with the office of Mr. William H. Hayes, State Inspector of Mines for approval of the portable diesel powered generator unit in the mines.

The field research trailer, which had been modified for sleeping, eating and laboratory space was taken to the Ambrosia Lake area and placed next to the hoist shed on September 9, 1969.

Mr. Wyrick had looked at likely experimental sites in the two mines and had an outline of the general airstream flow on the mine plots; the Section 23 mine seemed to offer the best possibilities.

Mr. Wyrick took John Doerges and Ed Powers along the proposed route

Figure 1. Counting equipment in mine location.

generally following the airstream route from the "8 East Vent Raise" (the air entry point) to the "68 East Exhaust Portal". Representative locations along the airstream were sampled. Figure 2 is a chart of that portion of the mine where sample stations were located. The following is a description of the sites:

Site 1. Bottom of the 8 East Vent Raise.

Outside air was being delivered through the vent raise to the mine at from 119,000 to 132,000 cubic feet per minute; the lower figure from mine data and the higher from a Dwyer pitot tube anemometer with corrections for barometric pressure, temperature and humidity. The air stream velocity was approximately sixty miles per hour. Site 1 is not on the map and is roughly a quarter of a mile in the direction of the arrow near the ordinate scale.

Site 2. 48 East Branch.

One branch had been blocked off by a bulkhead and a small car shop had been set up in the blocked off portion. This provided an ideal place to place the counting equipment (with lights!) and the samples could be taken in the main air stream. At this station the air stream had not passed through any mining areas so that any increase in radon and/or radon daughters can be attributed to seepage through the bulkheads and from the walls of the haulage drift.

Site 3. 85 East, 51 South; Access Air Door.

The ventilation air stream moves generally in the direction of increasing site numbers. Site 3 was chosen because it represents a point where the ventilation airflow recombines after splitting West of Site 2. An air access door leads to a mined out area immediately to the East of the station. Sites 4 and 5 are less than two hundred feet from Site 3 but approximately twenty to thirty feet lower in elevation.

Site 4. 90 East, 56 South; Mining Area.

A wagoner was hauling ore by the sample site and a miner was drilling near Site 5 during the swing shift. This location represented the oldest air in the mining area and exhausted through a vent raise into the mined out area East of Site 3 and

Figure 2. Diagram of portion of the 645 foot level of United Nuclear Homestake Partners Section 25 mine. Air sampling sites for aerosol analyses are indicated by numerals 1 through 7, as described in text.

subsequently into the 68 East Exhaust Portal without circulation into other mining areas.

Site 5. 87 East, 55 South; Mining Site.

An active mining area in relatively good ore for the Section 25 mine. The site was a typical mining area and was being actively mined.

Site 6. 73 East, 19 South; Buggy Barn.

Maintainence crews work nearby. Site was near the end of the ventilation distribution system. The site should give samples of the "oldest" air on the 645 level of the mine to which miners are regularly exposed.

Site 7. 66 East, 16 South; 68 East Exhaust Portal.

An exhaust fan on the surface pulls in excess of 100,000 cubic feet per minute of air out of the mine. The mine is therefore under a constant negative pressure and air that is delivered into some abandoned mined out areas from the ventilation stream eventually reaches the borehole and is exhausted to the surface. Air doors block the exhaust system from areas commonly occupied by miners. Miners spend a minimum of time (a few minutes/day) at this site. Analysis of mine atmosphere here should represent "well aged" air.

Experimental data were taken at each site over approximately six to eight hour periods. Site 1 on September 11th, Site 2 on September 12th, and Sites 3 through 7 on September 13th through 17th respectively. Data were taken at night during the last half of the swing shift and the first half of the graveyard shift (the mine is shut down during graveyard). These hours allowed us to collect data during active and passive mining periods and also during the transition between the two times. The following experimental data were recorded.

1. Barometric pressure.
2. Temperature.
3. Humidity.
4. Condensation Nuclei counts.

5. Radon samples (Lucas chambers).
6. Open filter air samples.
7. Diffusion tube air samples at several air flow rates.
8. Ventilation stream flow rates.
9. TLD air sample readings.

Analysis of data is in progress. Figure 3 is a graph of measured radon concentrations vs time at each of the seven sites. Figure 4 is a graph of measured working levels vs time for the seven sites. Table I is a summary of the radon daughter concentrations.

The open filters and diffusion tube filters were counted for RaA and RaC' alpha activity using silicone surface barrier detectors and collecting the counts on two one-hundred channel segments of a four-hundred channel analyzer. The counts under the RaA and RaC' peaks were integrated and appropriately recorded.

The resulting data have been punched on IBM cards and run on the diffusion tube analysis program written by D.E. McCurdy and D.F. Holleman. Results are presently being analyzed. The working levels graph of Figure 4 represents part of the results. A comprehensive report will be written on completion of the analysis.

The TLD samples were taken by Mr. Powers, under a research contract with the U.S. Public Health Service. Working levels from the TLD will be compared with the integrated working levels at each site from the multichannel analyzer in order to check calibrations.

VI. UPDATING OF COMPUTER PROGRAMS.

Three projects are in progress:

- 1) Listing all computer programs developed and/or used in the Radon Progeny Inhalation Study on computer tape. Collected data from experiments are also on tape. A backup tape is maintained in case of accidental loss due to machine or programming errors. The tape is maintained and modified by an UPDATE subroutine.

Figure 3. Radon concentrations at various sampling sites in United Nuclear Homestake Partners Section 25 mine by time of day.

Numerals indicate appropriate concentration scale for each sampling location as described in the text.

Figure 4. Measured working levels at various sampling sites in United Nuclear Homestake Partners Section 25 mine by time of day.

Numerals indicate appropriate working level scale for each sampling site. Standard deviations are not shown.

Table I. Radon Progeny Concentration Summary.

Site No.	Average Radon Concentration (pCi/l)	Standard Deviation of Rn Concentration (pCi/l)	Average Working Level	Standard Deviation of Working Level	Relative Activity of RaB : RaA	Relative Activity of RaC : RaA
1	12.79	1.23 (9.6%)				
2	211.22	22.3 (10.5%)	.523	.1415 (27.1%)	0.855 ± .484	0.732 ± .223
3	253.10	18.9 (7.5%)	.553	.087 (15.7%)	0.452 ± .198	0.318 ± .084
4	641.72	59.6 (9.3%)	1.698	.252 (14.8%)	0.392 ± .101	0.355 ± .065
5	435.26	32.2 (7.4%)	1.387	.182 (13.1%)	0.606 ± .142	0.464 ± .057
6	606.31	56.6 (9.3%)	.862	.124 (14.4%)	0.246 ± .077	0.195 ± .041
7	6300	435 (6.9%)	10.342	1.265 (12.2%)	0.377 ± .085	0.258 ± .031

Column 1 indicates site on map where measurements were taken. Column 2 is the average radon concentration at the site for all radon measurements taken. Column 3 is the standard deviation of the radon concentration. Columns 4 and 5 contain the average working levels and the standard deviations for each site. Columns 6 and 7 indicate the relative activity of RaB to RaA and RaC to RaA respectively. A ratio of 1.0 in both columns represent equilibrium conditions. The relatively high concentration of RaA to RaB and RaC yields higher working level values when measured by the spectroscopic method that which would be obtained by the Kusnet method (used by mine personnel).

2) The diffusion tube analysis program for radon daughters is being modified in four areas to give more versatility. Correction factors are being added to reduce standard deviations. Tests are being incorporated to determine if the experimental data are valid. All data from a given experiment will be read into an array at the beginning of the program allowing comparison between duplicate samples. The total program is being optimized to save computer time and to give alternate analysis and read-outs via subroutines.

3) A computer program is being written to test the aerodynamic size distributions of radioactive aerosols in uranium mines from the diffusion tube data. The analysis is based on the graph in Figure 5. The fraction of aerosols (F) penetrating a diffusion tube of a given length (Lcm) for different flow rates (Q) is plotted against the effective aerodynamic radius of the aerosol (r). From the diffusion tube experiments we can determine the fraction of radioactive aerosols which penetrate the diffusion tube between two different flow rates. This fraction can be represented by the integral equation

$$F = \int P(r) (F_1(r) - F_2(r)) dr$$

where

$F \equiv$ The difference in the two fractions of the number of aerosols that penetrate a diffusion tube.

$R_i \equiv$ The smallest effective aerodynamic aerosol radius in the distribution.

$R_f \equiv$ The largest effective aerodynamic aerosol radius in the distribution.

$P(r) \equiv$ A probability distribution as a function of aerosol radius.

$F_1(r) \equiv$ The fraction of aerosol at specified radius r penetrating the diffusion tube at the higher flow rate.

$F_2(r) \equiv$ The fraction of aerosol at specified radius r penetrating the diffusion tube at the lower flow rate.

Figure 5. Fraction (F) of aerosols which penetrate a diffusion tube as a function of effective aerodynamic radius (r) at seven flow rates.

Curve nearest ordinate is highest flow rate, curve most distant from ordinate represents lowest flow rate. Flow rates per diffusion tube are: 0.167, 0.133, 0.100, 0.080, 0.053, 0.027, and 0.013 liters/minute.

By assumming different aerosol distributions, (say normal or log normal), and employing numerical integration and approximation methods on the computer, one can determine the mean (or geometric mean) radius of the aerosol and the associated standard deviation of the population for the assumed distribution. The degree of agreement of the mean radius and standard deviation for several pairings of flow rates can be used as a measure to test the distribution assumed.

VII. DEVELOPMENT OF A SIZE DISCRIMINATING THERMAL PRECIPITATOR.

Thermal precipitators have been shown to segregate aerosol particles according to size, such that the average particle size precipitated increases with distance from the front edge nearest the intake towards the exit of the precipitator.

Figure 6 is a dimensionless perspective sketch of the thermal precipitator. Sides have not been included for reasons of clarity. The precipitator consists of two pie shaped metal plates, placed at an angle to each other so that the cross sectional area between the plates is constant. A temperature gradient is maintained between the two plates, the top having the higher temperature to prevent convection effects. The aerosol to be analyzed is introduced in a thin layer parallel to the base plate. Clean "carrier" gas is introduced above and below the aerosol layer at the same linear flow velocity as the aerosol. Flow rates are such that laminar flow is maintained.

The following mathematical derivation supports the premise that size discrimination of aerosols by thermal precipitation using a precipitator having the geometry described is feasible.

PERSPECTIVE SKETCH OF SIZE DISCRIMINATING THERMAL PRECIPITATOR

Figure 6.

Consider Figure 7 to represent a cross section of the aerosol chamber in the x-z plane.

Figure 7.

Definitions:

$h \equiv$ The vertical distance (altitude) between the top and bottom plates.

$P \equiv$ The point of intersection of the two plates extended.

$L \equiv$ The distance from the aerosol entrance to point P.

$x \equiv$ The distance from point P to base of altitude h.

$T_1 \equiv$ The temperature in degrees C of the top plate.

$T_2 \equiv$ The temperature in degrees C of the bottom plate.

$\Delta T \equiv T_1 - T_2$.

$\nabla T \equiv \Delta T/h$.

$v \equiv$ The average stream velocity of the aerosol and the diluting gas.

$\theta \equiv$ The angle formed by the two plates.

$d \equiv$ The distance the aerosol travels in time t.

$P \equiv$ The point of intersection of the two plates extended.

From the figure:

$$\tan \theta = h/x \quad (1)$$

$$d = L - x$$

$$d = vt \quad \text{which yields}$$

$$x = L - vt \quad (2)$$

from Eq. (1) and (2)

$$\nabla T = \Delta T/(x \tan \theta) \quad (3)$$

The force equation can be expressed by

$$F_N = F_T - F_R \quad \text{where}$$

$F_N \equiv$ The net force acting on a particle.

$F_T \equiv$ The thermophoretic force acting on the particle.

$F_R \equiv$ The resistance force of the media acting on the particle.

Equation (3) can be written

$$m \frac{d^2 y}{dt^2} = \frac{\pi p l r^2 T}{2T} - \frac{6\pi\eta(dy/dt)}{1 + A(1/r) + Q} \quad (4)$$

where

$m \equiv$ The mass of the particle (grams).

$p \equiv$ The pressure of the gas (dynes/cm²).

$l \equiv$ The mean free path of the gas molecules (cm).

$r \equiv$ The effective aerodynamic radius of the aerosol (cm).

$\eta \equiv$ The viscosity of the gas media (poises).

$A, Q, b \equiv$ Slip correction factor coefficients = 1.246, 0.42, and 0.87 respectively.

$$\text{Let } C = 1 + A(1/r) + Q(1/r) \exp(-br/l) \quad (5)$$

Equation (4) can be written (substituting from (1) and (5))

$$\frac{d^2 y}{dt^2} + \frac{6\pi\eta r}{mC} \frac{dy}{dt} = \frac{\pi p l r^2 \Delta T}{2mT \tan \theta} \cdot \frac{1}{x} \quad (6)$$

Let

$$\lambda = \frac{6\pi\eta r}{mC} = \frac{3 \cdot 6\pi\eta r}{4\pi r^3 p C} = \frac{9\eta}{2r^2 C} \quad (7)$$

$$\beta = \frac{\pi p l r^2 \Delta T}{2mT \tan \theta} \quad (8)$$

Substitution of (1) (7) and (8) in (6) yields

$$\frac{d^2 y}{dt^2} + \lambda \frac{dy}{dt} = \frac{\beta}{L-vt} \quad (9)$$

Introducing the integrating factor $e^{\lambda t}$ yields

$$e^{\lambda t} \frac{d^2 y}{dt^2} + \lambda e^{\lambda t} \frac{dy}{dt} = \frac{\beta e^{\lambda t}}{L-vt}$$

$$\frac{d(e^{\lambda t} (dy/dt))}{dt} = \frac{\beta e^{\lambda t}}{L-vt}$$

$$d(e^{\lambda t} (dy/dt)) = \beta \frac{e^{\lambda t}}{L-vt} dt \quad \text{or}$$

$$e^{\lambda t} \frac{dy}{dt} = \beta \int \frac{e^{\lambda t}}{L-vt} dt \quad (10)$$

Consider

$$e^{\lambda t} (L-vt)^{-1} dt = \frac{e^{\lambda t}}{\lambda(L-vt)} - \frac{v}{\lambda} \int \frac{e^{\lambda t}}{(L-vt)^2} \quad (11)$$

From integration by parts where

$$\begin{aligned} u &= (L-vt)^{-1} & dv &= e^{\lambda t} \\ du &= v (L-vt)^{-2} & v &= \frac{e^{\lambda t}}{\lambda} \end{aligned}$$

also

$$\frac{e^{\lambda t}}{(L-vt)^2} = \frac{e^{\lambda t}}{\lambda(L-vt)^2} - \frac{2v}{\lambda} \int \frac{e^{\lambda t}}{(L-vt)^3} \quad (12)$$

From integration by parts where

$$\begin{aligned} u &= (L-vt)^{-2} & dv &= e^{\lambda t} \\ du &= v (L-vt)^{-3} & v &= \frac{e^{\lambda t}}{\lambda} \end{aligned}$$

Substitution of (11) and (12) in (10) and repeated integration yields

$$\begin{aligned} e^{\lambda t} \frac{dy}{dt} &= \beta \left[\frac{e^{\lambda t}}{\lambda(L-vt)} - \frac{v e^{\lambda t}}{\lambda^2 (L-vt)^2} + \frac{2v^2 e^{\lambda t}}{\lambda^3 (L-vt)^3} + \dots \right. \\ &\quad \left. \frac{(-1)^{n-1} n! v^{n-1} e^{\lambda t}}{\lambda^n (L-vt)^n} + \dots \right] + C_1 \end{aligned}$$

where the first term in brackets is the zeroth term.

Rearranging 13

$$\frac{dy}{dt} = \frac{\beta}{\lambda} \sum_{n=0}^{\infty} \frac{(-1)^{n-1} n!}{(L-vt)^n} \frac{v^{n-1}}{\lambda} + C_1 e^{-\lambda t} \quad (14)$$

Integrating

$$y = \frac{-\beta}{\lambda v} \ln (L-vt) + \frac{\beta}{\lambda} \sum_{n=1}^{\infty} \frac{(-1)^n n! v^{n-1}}{\lambda^n (L-vt)^n} - \frac{C_1 e^{-\lambda t}}{\lambda} + C_2 \quad (15)$$

Now

$$\frac{\beta}{\lambda} = \frac{\pi p l r^2 \Delta T}{2 m T \tan \theta} \cdot \frac{m C}{6 \pi \eta r} = \frac{p l \Delta T C r}{12 T \eta \tan \theta} \quad (16)$$

By choice of the coordinate system $\Delta T < 0$ and all terms in (16) are constant except r , therefore one can write (15) as

$$y \sim r \left[\frac{1}{v} \ln (L - vt) + v \frac{(2\rho r^2 C)^2}{(9\eta(L - vt))^2} - 2v^2 \frac{(2\rho r^2 C)^3}{(9\eta(L - vt))^3} + \frac{3 \cdot 2v^3 (2\rho r^2 C)^4}{(9\eta(L - vt))^4} - \dots \right] - \frac{2\rho r^2 C}{9\eta} e^{-\lambda t} \quad (17)$$

where

$$\frac{1}{\lambda} = \frac{2}{9} \frac{\rho r^2 C}{\eta}$$

$$m = \frac{4}{3} \pi r^3 \rho$$

$\rho \equiv$ The particle density.

Equation (17) can be made convergent by proper choice of the parameters L , v , and t , particularly for small r .

By plotting curves of y as a function of r at different parameter values of L , v , and t one should be able to determine optimum dimensions and operating conditions for the thermal precipitator. A computer program is being written to check the above.

VIII. DEVELOPMENT OF A DIRECT READING, WORKING LEVEL SURVEY METER

A continuing need exists for a portable instrument capable of making direct measurements of radon progeny concentrations in uranium mine atmospheres in units of working levels. The basic criteria for such an instrument would be as follows:

- 1) It should be adaptable to easy use by mine personnel, i.e. it should be rugged, light weight and extremely simple to operate.
- 2) The readout should be directly in units of working levels, perhaps within ± 20 percent of the true value with little possibility for errors due to incorrect use of the instrument.
- 3) Measurements of working level concentrations should be independent of the degree of equilibrium of the radon progeny in the atmosphere.
- 4) The background rate should be low and stable.

A basic concept for a direct reading, working level survey meter was mentioned in our renewal proposal submitted in February 1968; however, progress toward completion of the prototype unit has been repeatedly delayed because of emphasis placed on other aspects of the research project. Since the need for such an instrument has not diminished during the past two years the efforts to produce a prototype have been revived during the past few months.

In keeping with the criteria for the instrument as stated above a few basic decisions were made regarding the construction of a prototype:

1. Air sampling will be accomplished by means of a hand operated pump which would draw a fixed volume of air through a filter for each stroke of the pump handle. This concept eliminates the need for the larger battery capacity which would be required by a motor driven pump and also eliminates the need for flow metering devices, timing of sampling periods, and precise calibration of air flow rates. At present it appears that a hand

pump which will draw 1 liter per stroke will be adequate.

2. Radioactivity counting will be accomplished by means of a single, solid-state surface-barrier detector. Detectors for both alpha and beta activities were not considered necessary or advisable; a beta particle detector by itself would not be adequate to indicate working levels in highly ventilated mine atmospheres where the effective age of the radon would be less than ten minutes.
3. Aerosol samples will be collected on a strip of membrane filter tape which will be advanced a fixed distance after each sample is drawn to place the sampled activity under the detector. This method was selected in order to eliminate the inconveniences associated with the handling of many small filter papers by the user of the instrument, and also to reduce the possibility of contamination of the detector surface.
4. A simple discriminator will be used to separate the alpha particle pulses from radium-A and radium-C', after which these pulses would be recombined electronically with appropriate weighting factors to give a readout equivalent to actual working levels in the atmosphere.

The relative contribution of the individual radon daughter products to the working level concentration are shown in Figure 8 as a function of the effective "age" of radon. From the curves in Figure 8 it is apparent that for radon progeny in complete equilibrium with the radon, or for atmospheres at least approaching equilibrium conditions, a measurement of any one of the three daughter products might give a reasonable estimate of total working level concentrations. As the mine operators attempt to reduce exposures by increasing ventilation, however, the problem becomes more and more difficult, owing to the increasing non-equilibrium condition of the radon progeny mixture. For "three minute air", approximately 85 percent of the working level concentration is due to radium-A and there is essentially no contribution from radium-C and C'. An empirical approach to the problem suggested that multiplication of the radium C' pulses from a detector by an appropriate weighting factor and adding these to the pulses from radium-A might produce an instrument response proportional to working levels but not dependent

Figure 8. Contribution of individual radon daughter products to working level concentrations as a function of the effective "age" of radon.

upon equilibrium or non-equilibrium conditions. Table II contains the calculated response of a direct reading survey meter for various atmospheres containing a one working level concentration.

Predicted response of the survey meter in counts per minute per working level is shown in Figure 9 as a function of an arbitrary equilibrium index, i.e. pCi of radium-A per working level liter. It can be seen that for a radium-C' multiplication factor of 10 a theoretical error of less than 10 percent might be achieved over a useful range from "three minute air" down to relatively stagnant air. An actual response that even approximated this theoretical response would be more than adequate for a portable survey instrument.

The general concept for the direct reading, working level survey meter is illustrated in Figure 10. The cylindrical instrument housing serves as the casing of the hand operated pump which draws one liter per stroke. The membrane filter tape would be contained on a roll inside the housing and would be advanced by pressing a release button and pulling on the end of the tape until the stop position was reached. The excess tape could be torn off against a serrated cutting edge located near the outer surface of the casing. The surface barrier detector and the electronic circuitry would be located in the end of the casing which would also house the readout meter.

To make a measurement, the pump plunger would be moved one full stroke to draw one liter of air through the filter tape. The tape would then be advanced to the stop placing the exposed portion of the tape directly beneath the detector. The counting sequence would then be initiated and, in one version of the instrument, a direct indication of working levels would be read on the meter. In the second version of the instrument, there would be a predetermined delay with the final measurement in working levels being displayed at the end of the counting period. In the event that the working level concentration was too low to be measured a second sample could be taken by making more than one stroke of the pump. The final measurement would then be obtained by dividing the output reading by the number of strokes taken on the pump.

As previously suggested we are working on two different approaches to the

Table II. Calculated response of a direct reading, working level survey meter for various 1 WL atmospheres.

Exposure Atmosphere			Meter Response Calculations*		
Nuclide	pCi/1	dpm/1	Detector α/m	Meter Count Rate	
				RaA + 8 RaC'	RaA + 10 RaC'
1. Equilibrium:					
RaA	100	220	52		
RaC'	100	220	55	492	602
2. Poor ventilation:					
RaA	275	600	141		
RaC'	50	110	27	357	411
3. Average ventilation:					
RaA	400	880	208		
RaC'	20	66	16	336	368
4. "Three minute air":					
RaA	810	1780	420		
RaC'	1.3	3	1	428	430
5. Complete disequilibrium:					
RaA	1000	2200	520		
RaC'	0	0	0	520	520

* Assumptions:

- Sampled volume = 1 liter
- Alpha counting yield = 0.25
- Decay time = 15 sec.

Figure 9. Survey meter response (cpm per WL) as a function of the degree of equilibrium of airborne radon progeny (pCi RaA per WL-liter).

For the ventilation conditions generally encountered in uranium mines, the concentration of RaA per WL-liter would be in the range of approximately 200 to 800 pCi. The meter response for the two circuit weighting factors would be (cpm/WL-liter):

$$\text{RaA} + 8 \text{ RaC}': 380 \pm 45 (\pm 12\%)$$

$$\text{RaA} + 10 \text{ RaC}': 400 \pm 32 (\pm 8\%)$$

Figure 10. Direct reading, working level survey meter (conceptual diagram).

electronic circuitry for the survey instrument. The first concept illustrated in Figure 11 is a simple count rate meter circuit with pulses from the detector passed through an amplifier and then into two separate signal channels. The first channel takes all the alpha pulses from both radium-A and radium-C' and sends them to the mixing circuit. The second channel contains a discriminator which allows only the radium-C' to pass; these pulses are then multiplied by a factor of approximately eight or ten before entering the mixer. The count rate meter output would then disclose a count rate calibrated to read in working level units. If one assumes that a liter of air can be drawn through a filter paper in a period of 15 seconds and that the surface-barrier detector would have a counting yield of approximately 25 percent, one arrives at a predicted count rate of 400 ± 32 counts per minute per working level.

The second concept being developed for the electronic circuitry involves an eight stage binary scaler (Figure 12). The detector and amplifier would be the same as in the previous circuit. The first channel of pulse information containing both radium-A and radium-C' pulses would be connected to the first stage of the scaler. The radium-C' pulses coming from the discriminator would be fed into the third stage of the binary scaler thus giving an effective multiplication factor of eight. A digital display output would be controlled by a countdown circuit, such that when the counting sequence was initiated, the display unit would count backwards from a high value towards zero. The last stage of the binary scaler (perhaps with a gain of 256) would provide a signal to stop the countdown circuit. The digital display output would hold at the reading remaining at the time of the stop signal, thus indicating working levels directly.

Both of the counting circuits just described are being assembled in bread-board form at the present time. A substantial amount of test and calibration work must yet be done before the suitability of either circuit can be established. We are extremely hopeful, however, that one or both of these circuits can be made to give actual readings in working levels.

Theoretical response:

400 ± 32 counts/min per WL-liter

Figure 11. Count rate meter circuit diagram for direct reading WL survey meter.

Figure 12. Count down scaler circuit diagram for direct reading WL survey meter.

BIBLIOGRAPHY

1. Holloman, Dan F., Radiation dosimetry for the respiratory tract of uranium miners, Special Report on U.S. Atomic Energy Commission Contract No. AT(11-1)-1500, Department of Radiology and Radiation Biology, Colorado State University, Fort Collins, Colo. (Dec., 1968)
2. Gormley, P. G. and M. Kennedy. Diffusion from a stream flowing through a cylindrical tube. Proc. of the Royal Irish Academy, Vol. 52, Sect. A:1630169. (1949)
3. Pollak and A. L. Metnieks, On the determination of the diffusion coefficient of heterogenous aerosols by the dynamic method, *Geofisica pura e applicata*: Vol. 34, 183-190 (1957).
4. Fuchs, N. A., I. B. Stechkind and V. I. Starosselskii, On the determination of particle size distribution in poly disperse aerosols by the diffusion method. *Brit. J. of Appl. Phys.*: Vol. 13, 280-281. (1962)
5. Fuchs, N. A. , *The Mechanics of Aerosols*. Pergamon Press, Oxford, London, Edinburgh, New York, Paris, Frankfurt. 1964
6. Walkerhorst, W. Beitr. Silikoseforsh., Heft 18. (Referred to by Fuchs (5) pg. 65.) (1952)
7. McCurdy, David E., Thermoluminescent dosimetry for personnel monitoring of uranium miners. Special Report on U.S. Atomic Energy Contract No. AT(11-1)-1500. Dept. of Radiology and Radiation Biology, Colorado State University, Fort Collins, Colo. (June 1969)

RADON (222) DECAY SCHEME