RMIS View/Print Document Cover Sheet

This document was retrieved from the Boeing ISEARCH System.

Accession #: D196054796

Document #: SD-WM-RPT-124

Title/Desc:

VAPOR & GAS SAMPLING OF SST 241BY111 USING VAPOR
SAMPLING SYSTEM

2. To: (Receiving Organization) DISTRIBUTION	3. From: (Originating Organization) SAMPLING AND MOBILE LABS OM624	4. Related EDT No.: N/A
5. Proj./Prog./Dept./Div.: VAPOR/HTS/SAMPLING AND MOBILE LABS	6. Cog. Engr.: RICKY MAHON 3-7437	7. Purchase Order No.: N/A
8. Originator Remarks: 241-BY-111, SAMPLING USING THE VAPOR SAMPLING SYSTEM		9. Equip./Component No.: N/A
		10. System/Bldg./Facility: M0292/WSCF
11. Receiver Remarks:		12. Major Assm. Dwg. No.: N/A
		13. Permit/Permit Application No.: N/A
		14. Required Response Date: 17 JULY 95

15. DATA TRANSMITTED					(F)	(G)	(H)	(I)
(A) Item No.	(B) Document/Drawing No.	(C) Sheet No.	(D) Rev. No.	(E) Title or Description of Data Transmitted	Approval Designator	Reason for Trans- mittal	Origina- tor Dispo- sition	Receiv- er Dispo- sition
1	WCH-SD-WM-RPT-124	ALL	0	VAPOR AND GAS SAMPLING OF SINGLE- SHELL TANK 241-BY-111 USING THE VAPOR SAMPLING SYSTEM	Q	2	1	

16. KEY					
Approval Designator (F)		Reason for Transmittal (G)		Disposition (H) & (I)	
E, S, Q, D or N/A (see WHC-CM-3-5, Sec.12.7)		1. Approval 2. Release 3. Information	4. Review 5. Post-Review 6. Dist. (Receipt Acknow. Required)	1. Approved 2. Approved w/comment 3. Disapproved w/comment	4. Reviewed no/comment 5. Reviewed w/comment 6. Receipt acknowledged

17. SIGNATURE/DISTRIBUTION (See Approval Designator for required signatures)										(G)	(H)
Reason	Disp.	(J) Name	(K) Signature	(L) Date	(M) MSIN	(J) Name	(K) Signature	(L) Date	(M) MSIN	Reason	Disp.
2	1	Cog.Eng. R.D. MAHON	<i>R.D. Mahon</i>	8/22/95	S3-27	RESOURCE CENTER			R2-12	3	
2	1	Cog. Mgr. J.D. DORIAN	<i>J.D. Dorian</i>	8/22/95	S3-27						
2	1	QA C.J. STEPHAN	<i>C.J. STEPHAN</i>	8/22/95	H4-16	Central Files			A3-88		
3		DAVE BRATZEL			S7-21						
3		JOHN DEICHMAN			T6-03						
3		RAMPUR VISWANATH			S3-90						
3		D.S.I.			A3-36						

18. GLENN S. CAPRIO <i>Glenn S. Caprio</i> 7/20/95 Signature of EDT Originator	19. _____ Authorized Representative Date for Receiving Organization	20. JOHN J. DORIAN <i>John J. Dorian</i> 8/24/95 Significant Manager Date	21. DOE APPROVAL (if required) Ctrl. No. ☐ Approved ☐ Approved w/comments ☐ Disapproved w/comments
--	--	--	---

THIS PAGE INTENTIONALLY
LEFT BLANK

RELEASE AUTHORIZATION

Document Number: WHC-SD-WM-RPT-124, REV 0

Document Title: Vapor and Gas Sampling of Single-Shell Tank 241-BY-111 Using the Vapor Sampling System

Release Date: 9/7/95

**This document was reviewed following the
procedures described in WHC-CM-3-4 and is:**

APPROVED FOR PUBLIC RELEASE

WHC Information Release Administration Specialist:

Kara M. Broz

TRADEMARK DISCLAIMER. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

This report has been reproduced from the best available copy. Available in paper copy. Printed in the United States of America. To obtain copies of this report, contact:

Westinghouse Hanford Company - Document Control Services
P.O. Box 1970, Mailstop H6-08, Richland, WA 99352
Telephone: (509) 372-2420; Fax: (509) 376-4989

THIS PAGE INTENTIONALLY
LEFT BLANK

SUPPORTING DOCUMENT

1. Total Pages 34

2. Title

Vapor and Gas Sampling of Single-Shell Tank 241-BY-111 Using the Vapor Sampling System

3. Number

WHC-SD-WM-RPT-124

4. Rev No.

0

5. Key Words

241-BY-111
VSS
SUMMA
TST
Sorbent
Tank

6. Author

Name: G.S. Caprio

Signature

Organization/Charge Code OM624/N4131

7. Abstract

This document presents sampling data resulting from the November 16, 1994, sampling of SST 241-BY-111 Using the Vapor Sampling System.

8. RELEASE STAMP

THIS PAGE INTENTIONALLY
LEFT BLANK

APPROVALS

Document title: Vapor and Gas Sampling of Single-Shell Tank 241-BY-111
Using the Vapor Sampling System

Approved by:

R. A. Westberg
R. A. Westberg, Field Scientist
Vapor Sampling Project
Sampling and Mobile Laboratories

6/5/95
Date

Approved by:

Rick Mahon
R. D. Mahon, Project Lead
Vapor Sampling Project
Sampling and Mobile Laboratories

22 JUL 95
Date

Approved by:

D. L. Edwards
D. L. Edwards, Technical Lead
Vapor Sampling Project
Sampling and Mobile Laboratories

6/5/95
Date

Approved by:

R. S. Viswanath
R. S. Viswanath, Principal Scientist
Vapor Sampling Project
Special Analytical Studies

6-5-95
Date

THIS PAGE INTENTIONALLY
LEFT BLANK

CONTENTS

1.0	SCOPE	1
2.0	SAMPLING EQUIPMENT DESCRIPTION	1
2.1	VAPOR SAMPLING SYSTEM	1
3.0	SAMPLING EVENT DESCRIPTION	3
3.1	SPECIFICATIONS	3
3.2	OPERATIONS AND SAMPLING PERSONNEL	3
3.3	AMBIENT CONDITIONS	3
3.4	INDUSTRIAL HYGIENE FIELD RESULTS	4
3.5	SAMPLE COLLECTION	4
3.6	FIELD GC/FID RESULTS	5
3.7	RADIATION SCREENING	5
4.0	SAMPLE CHAIN OF CUSTODY: RECEIPT, STORAGE, AND SHIPMENT	7
5.0	QUALITY ASSURANCE AND CONTROLS	9
5.1	VAPOR SAMPLING SYSTEM CLEANING	9
5.2	INSTRUMENT CALIBRATION	9
5.3	BLANK SAMPLES	10
6.0	ANOMALIES	10
7.0	REFERENCES	11
APPENDICES		
A	SAMPLE LOG SHEETS	A-1
B	AMBIENT CONDITIONS	B-1
C	CHAIN-OF-CUSTODY FORMS	C-1

LIST OF TABLES

1	Flow Control Calibration	2
2	TOC Results	5
3	Radionuclide Analysis Results	6
4	Oak Ridge National Laboratory Samples	7
5	Pacific Northwest Laboratory Samples	8
6	Oregon Graduate Institute Samples	8
7	Calibration Data	9
8	Blank Samples	10

LIST OF TERMS

COC	Chain Of Custody
DOT	U.S. Department of Transportation
GC/FID	Gas Chromatograph/Flame Ionization Detector
GEA	Gamma Energy Analysis
HEPA	High-Efficiency Particulate Air (filter)
NH ₃	Ammonia
NO ₂	Nitrogen Dioxide
NO	Nitric Oxide
H ₂ O	Water Vapor
OGIST	Oregon Graduate Institute of Science and Technology
OPC	Offsite Property Control
ORNL	Oak Ridge National Laboratory
PNL	Pacific Northwest Laboratory
SML	Sampling and Mobile Laboratories
SST	Single-Shell Tank
TCP	Tank Characterization Plan
team	SML Vapor Team
TOC	Total Organic Carbon
TST	Triple Sorbent Trap
VSS	Vapor Sampling System
WHC	Westinghouse Hanford Company

This page intentionally left blank.

VAPOR AND GAS SAMPLING OF SINGLE-SHELL TANK 241-BY-111 USING THE VAPOR SAMPLING SYSTEM

1.0 SCOPE

The Vapor Issue Resolution Program tasked the Sampling and Mobile Laboratories (SML) to collect representative headspace samples from Hanford Site single-shell tank (SST) 241-BY-111. This document presents sampling data resulting from the November 16, 1994, sampling of SST 241-BY-111.

2.0 SAMPLING EQUIPMENT DESCRIPTION

2.1 VAPOR SAMPLING SYSTEM

The SML vapor team (the team) used the vapor sampling system (VSS) to collect representative samples of the air, gases, and vapors from the headspace of SST 241-BY-111 on November 16, 1994. Mahon et al. (1994) describes in detail the VSS, its performance, and its operation. The team used the VSS to collect sorbent traps and SUMMA¹ canister headspace samples from SST 241-BY-111; subsequently, the team sent these samples to analytical laboratories for analysis.

The VSS comprises a mobile laboratory, a hot-water-jacketed stainless steel probe that is inserted into the tank headspace, and stainless steel transfer tubing that connects the mobile laboratory to the probe. A vacuum pump draws sample vapor from the tank headspace and through all transfer tubing and the sampling manifold. Electrically activated, pneumatically actuated valves direct sample flow within the VSS. Instrumentation housed in the mobile laboratory monitors and controls system temperature, monitors absolute and differential system pressure, meters and controls mass flow, and monitors sample vapor using a gas chromatograph/flame ionization detector (GC/FID).

A key feature of the VSS is its use of heated transfer tubing and a heated sampling manifold. Maintaining the system temperature at an electronically controlled elevated temperature prevents vapor condensation and reduces vapor adsorption on surfaces exposed to sample vapor. Mahon et al. (1994) describes various tests and observations that indicate the VSS sample transfer efficiency is consistently high.

Sorbent trap samples are collected at the sorbent trap station of the sampling manifold. Sorbent traps are pencil-size stainless steel or glass tubes that contain vapor-adsorbing media. A known amount of sample vapor is passed through the tube, which traps (by adsorption) virtually all the target analytes. The sorbent trap is then sent to an analytical laboratory and

¹SUMMA is a registered trademark of Molectrics, Inc., Cleveland, Ohio.

analyzed. The concentration of analytes in the vapor sampled is calculated from the quantity of analyte found in the sorbent media and the volume of vapor passed through the sorbent trap.

The sorbent trap station uses highly accurate mass flow controllers to measure and control the flow rate of sample vapor through the sorbent traps. The controllers are located downstream of the sorbent trap station and the in-line driers, which remove water vapor from the vapor before the vapor is metered. Errors associated with the mass flow controllers were determined by the Westinghouse Hanford Company (WHC) Standards Laboratory before the SST 241-BY-111 sampling (see Table 1).

The VSS is also equipped with a gas chromatograph (GC). The Hewlett Packard 5890 Series II GC is equipped with a flame-ionization detector (FID), 1 mL sample loop, 10 port injection valve, 2 meter chromatographic column, programmable oven, portable computer, and HP-Chemstation software to control the GC. The GC is plumbed to directly transfer the sample from the VSS manifold to the sample loop. After the sample is transferred into the sample loop and reaches equilibrium, the run is initiated manually. HP-Chemstation software activates the 10 port valve to transfer the sample from the sample loop to the column. The sample passes through the column and the FID generates a signal for total organic carbon (TOC). All the data is then transmitted to the computer where it is stored for further analysis.

The GC is equipped with a HP-5 column which is 2 meters long, 0.25 mm inside diameter, and a 0.25 μ m phenyl methyl silicone phase. The GC oven is programmed to heat from 50 °C to 270 °C at a rate approximately 70 °C per minute. Helium is the carrier gas, air and hydrogen the combustion gases, and nitrogen the make-up gas.

The gas chromatograph/flame ionization detector (GC/FID) is configured to quantitatively estimate concentrations of TOC. The GC/FID confirms sampling system cleanliness, sampling manifold performance, TOC of ambient air background levels during sampling, and TOC of tank vapor samples. The system is multi-point calibrated and has displayed a high degree of stability over a period of months. For further details, see Section 3.6, Field GC/FID Results. Flow rates and the duration of flow are specified by the analytical laboratories that supply and analyze the sorbent traps.

Table 1. Flow Control Calibration

Flow-indicating control valve	Flow (stdcm ³ /min)	Error (%)
1	200	± 0.15
3	200	± 0.30

The sampling manifold also has a station for sampling vapor with evacuated SUMMA[™] canisters. SUMMA[™] canisters are stainless steel vessels with internal surfaces that have been prepared by the SUMMA[™] process, which passivates active sites on the canister walls to minimize adsorption of gases and vapors. An analytical laboratory must clean and evacuate SUMMA[™] canisters before use.

The evacuated canister is filled with sample vapor through a valve, which is then closed to seal the sample inside. SUMMA™ canisters essentially allow collection and transfer of whole-air samples from location to an analytical laboratory where the sample is analyzed. The standard capacity of each SUMMA™ canister is 6 liters.

Prior to sampling a GC/FID confirms that the VSS is free of hydrocarbon contamination, to the limit of detection. The GC/FID is part of the VSS and can be used during a sampling event to periodically monitor organic vapor concentrations in the vapor being sampled.

3.0 SAMPLING EVENT DESCRIPTION

3.1 SPECIFICATIONS

The Vapor Issue Resolution Program specifies sampling requirements in WHC-SD-WM-TP-280, *Tank 241-BY-111 Tank Characterization Plan* (Homi 1995). The TCP also specifies the types and number of samples to be collected, flow rates, and durations. These key sampling parameters are summarized on the sample log sheets in Appendix A. In addition to the sample log sheets, checksheets for each individual sample help ensure correct sampling procedures. SML retains these documents in the project file. This sample event's project-specific number is S4-083.

3.2 OPERATIONS AND SAMPLING PERSONNEL

E. A. Tampos was the Tank Farm Operations person-in-charge. The other SML vapor team members included:

- E. A. Johnson, Field Scientist
- R. A. Westberg, Field Scientist
- R. L. Ragan, Field Scientist
- T. B. Utecht, Trailer Support.

The VSS was set up at SST 241-BY-111 on November 15, 1994 and was allowed to warm up. Sampling began shortly after 12:00 p.m. on November 16, 1994, and was completed by 4:00 p.m. the same day.

3.3 AMBIENT CONDITIONS

The weather the day of the sampling event, November 16, 1994, was cool and cloudy with light winds. Graphs of ambient temperatures and pressures taken at the Hanford Meteorological Station, which is about 4 miles due west of 241-BY tank farm, are provided in Appendix B.

3.4 INDUSTRIAL HYGIENE FIELD RESULTS

Before connecting to SST 241-BY-103, an industrial hygiene technician first field tested tank vapors. The technician purged the vapor probe sample tube for 5 minutes and then field measured vapor stream contents using a combustible gas indicator (CGI) and organic vapor meter (OVM). The technician verbally reported an LEL of 0.0% and a TOC content of 2.0 ppm using the OVM. The technician also used a Dräger tube to estimate the ammonia content of the tank as 45 ppm.

3.5 SAMPLE COLLECTION

The hot-water-jacketed sampling probe was located in Riser 12A of SST 241-BY-111. The probe length, from the sample inlet to the top of the riser flange, was 7.3 meters.

All zones of the VSS were heated to 60 °C during setup of the VSS at SST 241-BY-111 on November 15, 1994. The team stabilized the VSS temperature zones by 10:00 a.m. on November 16, 1994, and the system was ready to collect samples. Measured according to the VSS operating procedure, the pressure and temperature of SST 241-BY-111 were 976 mbar (732 torr) and 24 °C, respectively. The sample log sheets (Appendix A) provide a complete chronology of the sample event including start and end times, flow rates, volumes, and specific sample identifiers.

Approximately 19 hours after heating began, the team purged the sampling manifold with ambient air for 30 minutes. The team collected two SUMMA[™] canister samples of ambient air, one manually 10 meters upwind of the VSS connection with SST 241-BY-111, and one using the VSS sampling manifold. The former was collected to establish background levels of trace organic vapors, and the latter was collected to establish the cleanliness of the sampling manifold.

A leak check of the VSS sampling manifold and transfer tubing (up to the connection to the in-tank sampling probe) was performed. Leakage resulted in an increase of 8.8 mbar (6.6 torr) rise in pressure over the 15 minute test. Given a system volume of not more than 10 L, this pressure rise corresponds to a leak rate of approximately 19.9 mL/min at a 294 mbar (220.7 torr) system pressure. This leak rate was estimated for average SUMMA[™], TST, and sorbent sampling pressures. It was found that for the SUMMA[™] canisters, dilution by ambient air was approximately 0.2%, for TST traps sampled at 200 mL/min the dilution was approximately 0.8% and for sorbent traps the dilution was approximately 0.8%.

The VSS was purged with sample vapor from SST 241-BY-111 for 30 minutes at a total flow rate of 5.45 L/min. This purge was performed to flush ambient air from the system and saturate the system's active adsorption sites. Because the volume of transfer tubing and the sampling manifold upstream of the sampling devices is estimated to be no more than 10 L, this purge provided about 16.4 air turnovers in the system.

Three analytical laboratories provided sample media. Oak Ridge National Laboratory (ORNL) provided triple sorbent traps (TSTs) for organic vapors;

Pacific Northwest Laboratory (PNL) provided SUMMA™ canisters and sorbent traps for organic vapors, ammonia (NH₃), nitrogen dioxide (NO₂), nitric oxide (NO), and water vapor (H₂O); and the Oregon Graduate Institute of Science and Technology (OGIST) provided SUMMA™ canisters.

3.6 FIELD GC/FID RESULTS

The GC was single point calibrated on the day of sampling using a 19.0 ppmC Ethane standard by Scott Specialty Gas. The standard is an E.P.A. Protocol Gas that is +/- 2% NIST traceable. Table 2 displays the Field GC/FID results from the sampling of SST 241-BY-111.

Table 2. TOC Results.

Number of Runs	Description	Average TOC Concentration (ppmC)	% Standard Deviation
3	Ethane	19.00	1.09
3	Ambient	2.94	10.04
3	Tank Vapor	7.97	3.77

3.7 RADIATION SCREENING

Samples are unconditionally released from the SST farm in accordance with DI-33300-93-004, *Release of Vapor Sampling Equipment* (WHC 1993b). Radiological screening results are used to determine (1) if the samples must be shipped as radioactive or nonradioactive in accordance with U.S. Department of Transportation (DOT) regulations and (2) if the samples meet the laboratory acceptance criteria.

The DOT limits for shipping a nonradioactive sample are 200 pCi/g of beta-gamma activity and 60 pCi/g of alpha activity. Samples exceeding these DOT limits may be shipped as radioactive material, if the samples do not exceed the following laboratory acceptance criteria:

- OGIST: The laboratory has supplied no acceptance criteria for alpha-beta-gamma activity. Notify the laboratory if radioactive levels exceed the DOT limits specified above, or if >50 pCi/g tritium are found.
- ORNL: Beta-gamma activity <450 pCi/g of sample media. Alpha activity <135 pCi/g.
- PNL: Beta-gamma activity <400 pCi/g of sample media. Alpha activity <100 pCi/g of sample media.

To protect the sampling manifold and all sampling devices from radioactive particulates, all sample vapor for the November 16, 1994, SST 241-BY-111 vapor

sampling event was drawn through a series of four glass-fiber high-efficiency particulate air (HEPA) filters placed upstream of the sampling manifold. These four filters were in place any time tank headspace vapor, gases, and vapors were flowing through the system. When sampling was complete, the filters were removed and assigned unique sample identifiers. All four HEPA filters were submitted to Laboratory 222-S for total alpha, total beta, and gamma energy analysis (GEA). The HEPA filter closest to the sampling manifold was analyzed to determine if the samples met DOT shipping criteria and laboratory acceptance criteria. SST 241-BY-111 filter sample results are shown in Table 2. SML scientists use the activity results from Table 3 to calculate pCi/g of sample media. SML maintains this information in the project-specific file.

A water vapor sample was also collected through one of the sorbent station ports and analyzed for tritiated water. SST 241-BY-111 tritium sample results are shown in Table 3. The results in Table 3 indicate that the samples collected from SST 241-BY-111 met the laboratory acceptance criteria and the DOT definition of a nonradioactive shipment.

Table 3. Radionuclide Analysis Results.

Filter	Sample Identifier	Activity Results ^a (pCi/sample)	Activity ^b (pCi/L of tank gas)
Upstream HEPA filter (box)	S4083-A25.0U1	Total Alpha = 13600 Total Beta = 23100 GEA = 2800 (²⁰⁸ Tl) GEA = 8590 (²¹² Bi) GEA = 8850 (²¹² Pb)	= 34.78 = 59.08 = 7.16 = 21.97 = 22.63
Downstream HEPA filter (box)	S4083-A26.0D1	Total Alpha = <.972 Total Beta = 5.08 GEA = <detectable	= <0.002 = 0.01 = <detectable
Upstream HEPA filter (VSS)	S4083-A27.0U2	Total Alpha = .841 Total Beta = 3.01 GEA = <detectable	= 0.002 = 0.008 = <detectable
Downstream HEPA filter (VSS)	S4083-A28.0D2	Total Alpha = <1.01 Total Beta = <2.85 GEA = <detectable	= <0.003 = <0.007 = <detectable
Tritium trap	S4083-A03.0T1	Total activity = <50.0	= <50.0 ^c

NOTES:

The samples are nonradioactive. These results were evaluated against laboratory acceptance criteria and DOT limits (see Section 3.5).

^aAll less than (<) values represent the minimum detection limits at Laboratory 222-S.

^bNumbers based on an approximation of the total volume of tank vapor through the HEPA filters. Appendix A and the sample checksheets were used to estimate a total flow through the VSS of 391 L.

^cNumber is calculated using a total volume of 1 L passing through the tritium trap.

4.0 SAMPLE CHAIN OF CUSTODY: RECEIPT, STORAGE, AND SHIPMENT

All sorbent trains, sorbent tubes, and SUMMA™ canisters received by SML are stored in a custody locker maintained by SML. Sorbent trains and tubes were maintained at 4 ± 2 °C in a refrigeration unit. SUMMA™ canisters were stored in the same locked storage area, but were not refrigerated.

TSTs are supplied by ORNL. ORNL initiates the chain-of-custody (COC) forms. The samples are placed in 40-mL volatile organic analysis vials. Evidence tape is applied to the vials, which are then placed in a shipping container along with the relinquished COC forms and shipped to SML. TSTs are shipped to SML in a cooler containing "blue" ice.

The TSTs were received, inventoried against the COC forms, and the "received by" block on the COC form was signed by SML. They were maintained at 4 ± 2 °C before and after the sampling event in a refrigeration unit.

The TSTs were shipped to ORNL using offsite property control (OPC W95-0-0019). Table 4 lists the sample identifiers, sample types, and COC form numbers for all ORNL samples. The samples were delivered for transport on November 28, 1994, and shipped by Federal Express for next-day delivery.

Table 4. Oak Ridge National Laboratory Samples.

Sample Identifier	Sample Type	COC Number
S4083-A10.610	TST	006745
S4083-A11.644	TST	006745
S4083-A12.663	TST field blank	006745
S4083-A13.645	TST	006745
S4083-A14.648	TST	006745
S4083-A15.656	TST trip blank	006745

Sorbent trains and SUMMA™ canister sampling devices were supplied with COC forms by PNL. Sampling devices were picked up from PNL by SML and transported in a government vehicle to a custody locker.

After sampling, the PNL sorbent traps and SUMMA™ canisters were transported by government vehicle directly to PNL and delivered to J. A. Edwards on November 28, 1994. Table 5 lists the sample identifiers, sample types, and COC form numbers for all PNL samples. The sorbent trains were maintained at 4 ± 2 °C before and after the sampling event in a refrigeration unit.

Table 5. Pacific Northwest Laboratory Samples.

Sample Identifier	Sample Type	COC Number
S4083-A01.223	Ambient upwind SUMMA™	008076
S4083-A02.224	Ambient SUMMA™ (VSS)	008076
S4083-A07.247	SUMMA™	008076
S4083-A08.248	SUMMA™	008076
S4083-A09.249	SUMMA™	008076
S4083-A16.91V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A17.92V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A18.93V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A19.94V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A20.95V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A21.96V	NH ₃ /NO ₂ /NO/H ₂ O sorbent	008077
S4083-A22.97V	NH ₃ /NO ₂ /NO/H ₂ O trip blank	008077
S4083-A23.98V	NH ₃ /NO ₂ /NO/H ₂ O trip blank	008077
S4083-A24.99V	NH ₃ /NO ₂ /NO/H ₂ O trip blank	008077

SUMMA™ canisters were also supplied by OGIST. The sample media is shipped to SML without a COC form. The sample media is inventoried at SML, where the COC is initiated.

The canisters were transported by government vehicle directly to PNL for storage and delivered to J. A. Edwards on November 28, 1994. Table 6 lists the sample identifiers, sample types, and COC form numbers for all OGIST samples.

Table 6. Oregon Graduate Institute Samples.

Sample Identifier	Sample Type	COC Number
S4083-A04.H209	SUMMA™	008076
S4083-A05.H232	SUMMA™	008076
S4083-A06.H233	SUMMA™	008076

From the time that samples are received by SML until they are shipped back to the analytical laboratory, COC is maintained by SML in accordance with LO-090-443/A0, *Chain-of-Custody for RCRA and CERCLA Protocol Samples* (WHC 1992). The final COCs from this sampling event are in Appendix C.

5.0 QUALITY ASSURANCE AND CONTROLS

5.1 VAPOR SAMPLING SYSTEM CLEANING

Immediately prior to sampling of SST 241-BY-111, the VSS manifold was heated to 60 °C and ambient air was purged through the system for 30 minutes. Throughout this purge, all pertinent system valves were actuated to release any contaminants that may have collected in the VSS valves themselves. After this purge, an ambient air sample was drawn through the VSS manifold and a GC/FID run was initiated. No significant level of contaminants was detected by the FID. A second sample confirmed the VSS was free of organic remnants down to ambient levels. Thus, the system was ready to sample 241-BY-111. For further details, refer to Appendix C of LO-080-450/A0 (WHC 1993c) and the project-specific file located with SML.

In compliance with the *Vapor Space Probe Cleaning Procedure* (WHC 1993), the team cleaned the hot-water-jacketed probe before it was installed in SST 241-BY-111. This procedure requires a solvent rinse of all internal probe surfaces with acetone and methanol to clear the sample line of possible contamination remaining from the construction process. The probes are then heated to 90 °C, and dry air is passed through the probe to evaporate the solvents. Before the probes are released, the team ensures, using a hand-held organic vapor meter, that the total organic concentration in the sample line is below 1 ppmv.

SUMMA™ canister ambient air samples were also collected to confirm that the VSS sampling manifold was free of trace organic contaminants (or to determine what contaminants were present). One SUMMA™ canister was filled with ambient air upwind of SST 241-BY-111, and a second canister was filled with ambient air via the VSS sampling manifold. The canisters were shipped to PNL for analysis.

5.2 INSTRUMENT CALIBRATION

Instruments located in the VSS are calibrated on an annual basis at the WHC Standards Laboratory. VSS instrumentation calibration data, maintained in files by SML, is summarized in Table 7. According to the calibration schedule shown in Table 7, all instrumentation was within its calibration period during the SST 241-BY-111 sampling event.

Table 7. Calibration Data.

System	Calibration Date	Expiration Date	WHC Standards Laboratory Code
Temperature Control System	04/21/94	04/21/95	804-67-74-009
Mass Flow System	04/29/94	04/29/95	804-28-03-001
Pressure System	05/03/94	05/03/95	804-67-89-001

5.3 BLANK SAMPLES

Trip blanks are samples that accompany the sample media from the point of generation through sample analysis. Analysis of trip blanks is used to assess cross-contamination of the sample media during transport and storage. Trip blanks are transported to the field with the sample collection media but remain unopened throughout the sampling event.

Spiked blanks are samples that are spiked with a known amount of analyte before shipment. These blanks are handled and analyzed with the other samples. Analysis of the spiked blanks is used to evaluate potential sample loss during shipment or storage.

Ambient blanks are samples of ambient air collected at the sampling location. Analysis of ambient blanks is used to assess contamination that may be present in the atmosphere surrounding the sampling system or in the system itself prior to sampling.

Field blanks are sampling devices prepared and handled in the same manner as the sampler, but tank gases are drawn through them.

Table 8 lists the blank samples collected during the sampling of SST 241-BY-111.

Table 8. Blank Samples.

Sample Identifier	Blank Type	Sample Media
S4083-A01.223	Ambient	SUMMA™
S4083-A02.224	Ambient through VSS	SUMMA™
S4083-A15.656	Trip	TST
S4083-A12.663	Field Blank	TST
S4083-A22.97V	Trip	NH ₃ /NO ₂ /NO/H ₂ O
S4083-A23.98V	Trip	NH ₃ /NO ₂ /NO/H ₂ O
S4083-A24.99V	Trip	NH ₃ /NO ₂ /NO/H ₂ O

6.0 ANOMALIES

All samples were collected in accordance with the TCP and LO-080-450/A0, *Collection of SUMMA Canisters and Sorbent Tube Samples Using the Vapor Sampling System* (1993a). No anomalies to the TCP or the operating procedure were noted during the sampling of SST 241-BY-111. Further details are available in the project-specific file maintained by SML.

7.0 REFERENCES

- 49 CFR 100-177, 1992, "Transportation," *Code of Federal Regulations*, as amended.
- Homi, C. S., 1995, *Tank 241-BY-111 Tank Characterization Plan*, WHC-SD-WM-TP-280, Westinghouse Hanford Company, Richland, Washington.
- Mahon, R. D., C. M. Jones, and M. S. Story, 1994 (draft), *Evaluation of the Capabilities and Use of the Vapor Sampling System for Tank Headspace Sampling and Characterization*, SD-WM-RPT-094, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1993a, *Collection of SUMMA Canisters and Sorbent Tube Samples Using the Vapor Sampling System (VSS)*, Procedure LO-080-450/A0, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1993b, *Release of Vapor Sampling Equipment*, Procedure DI-33300-93-004, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1993c, *Vapor Space Probe Cleaning Procedure*, Procedure LO-080-405/A0, Westinghouse Hanford Company, Richland, Washington.
- WHC, 1992, *Chain-of-Custody for RCRA and CERCLA Protocol Samples*, Procedure LO-090-443/A0, Westinghouse Hanford Company, Richland, Washington.

This page intentionally left blank.

APPENDIX A
SAMPLE LOG SHEETS

This page intentionally left blank.

Set up VSS (Section A)(Temperature set point = 60C)

Ensure HEPA filters are installed

Ensure connection to sample probe

System status check sheet. (Verify zones are to temp)

Tank Temperature 24.0 C

Date: 16 November, 1994

VSS Personnel:

E.A. Johnson, R. Ragan,

R. Westberg

Trailer Personnel:

T. Utecht

PNL		Desired		Desired		Desired		Actual		Actual		Actual	
OGI/ORNL	Port	Description	Flow	Duration	Total	Flow	Total	Flow	Start	End	Total	Flow	Total
Sample ID	Valve #		Rate	(min.)	Flow	Rate	Flow	Rate	Time	Time	Time	Flow	Flow
WHC Sample ID					SCCM	SCCM		SCCM					Liters
Group # 1													
Purge with ambient air for 30 min.													
4083- A01	223	AMBIENT #1 (PNL)	5450	30	163500				10:24	10:25			
GC Run #1													
4083- A02	224	AMBIENT #2 (PNL)		1					10:30				
	15								11:00	11:01			
Leak check (Appendix A) Leak Rate: 26.0 Torr/Hour													
Purge with tank air for 30 min.													
			5450	30	163500	5459			11:37	12:11	34	185.61	
Measure tank pressure 732.0 Torr (PE-1)													
4083- A03	OT1	T1664(222S	200	5	1000	210.50			12:27	12:32	5	1.00	
GC Run # 2 (Tank run #1)													
4083- A04	H209	H209		1					12:48	12:49	1	6.00	
4083- A05	H232	H232		1					12:53	12:54	1	6.00	
4083- A06	H233	H233		1					12:57	12:58	1	6.00	
4083- A07	247	247		1					13:13	13:14	1	6.00	
4083- A08	248	248		1					13:17	13:18	1	6.00	
4083- A09	249	249		1					13:21	13:22	1	6.00	
GC Run # 3 (Tank run #)													

Waste Tank
241-BY-111

WHC Sample ID	PNL OGI/ORNL Sample ID	Port Valve #	Description	Desired		Desired		Actual		Actual		Actual	
				Flow	Rate	Duration (min.)	Total Flow	Flow	Rate	Start Time	End Time	Total Time	Total Flow
4083- A10 .610	TST# 610	9	TST #1	200	200	10	2000	200.30	200.30	13:40	13:50	10	2.00
4083- A11 .644	TST# 644	10	TST #2	200	200	10	2000	200.30	200.30	13:40	13:50	10	2.00
4083- A12 .663	TST# 663		TST FIELD BLANK #1							13:46	13:47		
4083- A13 .645	TST# 645	8	TST # 3	200	200	10	2000	200.30	200.30	13:58	14:08	10	2.00
4083- A14 .648	TST# 648	10	TST # 4	200	200	10	2000	200.30	200.30	13:58	14:08	10	2.00
GC Run # 4 (Tank run #3)													
4083- A16 .91V	91V	9	NH3/NOx/H2O #1	200	200	15	3000	200.30	200.30	14:29	14:44	15	3.00
4083- A17 .92V	92V	10	NH3/NOx/H2O #2	200	200	15	3000	200.30	200.30	14:29	14:44	15	3.00
4083- A18 .93V	93V	8	NH3/NOx/H2O #3	200	200	15	3000	200.30	200.30	14:50	15:05	15	3.00
4083- A19 .94V	94V	10	NH3/NOx/H2O #4	200	200	15	3000	200.30	200.30	14:50	15:05	15	3.00
4083- A20 .95V	95V	9	NH3/NOx/H2O #5	200	200	15	3000	200.30	200.30	15:11	15:26	15	3.00
4083- A21 .96V	96V	10	NH3/NOx/H2O #6	200	200	15	3000	200.30	200.30	15:11	15:26	15	3.00
TOTAL TANK GAS USED DURING SAMPLING RUNS												63.04	

Waste Tank
241-BY-111

WHC Sample ID	PNL		Description	Desired	Desired	Desired	Actual	Actual	Actual	Actual	Actual
	OGI/ORNL	Port		Flow	Duration	Total	Flow	Start	End	Total	Total
	Sample ID	Valve #		Rate	(min.)	Flow	Rate	Time	Time	Time	Flow
4083- A25 .0U1	T1665(222S	25	Upstream HEPA(box)						15:45		
4083- A26 .0D1	T1666(222S	26	Downstream HEPA(box)						15:45		
4083- A27 .0U2	T1667(222S	27	Upstream HEPA (VSS)						15:45		
4083- A28 .0D2	T1668(222S	28	Downstream HEPA (VSS)						15:45		

Trip Blanks (DO NOT EXPOSE)

4083- A15 .656	TST# 656	15	TST TRIP #1
4083- A22 .97V	97V	22	NH3/NOx/H2O TRIP #1
4083- A23 .98V	98V	23	NH3/NOx/H2O TRIP #2
4083- A24 .99V	99V	24	NH3/NOx/H2O TRIP #3

A-5

WHC-SD-WM-RPT-124, Rev. 0

[illegible]

APPENDIX B
AMBIENT CONDITIONS

This page intentionally left blank.

WEATHER FOR VSS VAPOR SAMPLING OF BY-111

Time on 11/16/94

This page intentionally left blank.

APPENDIX C
CHAIN-OF-CUSTODY FORMS

This page intentionally left blank.

WHC-SD-WM-RPT-124, Rev. 0

Westinghouse
Hanford Company

CHAIN OF CUSTODY

ORNL-006745

Custody Form Initiator Amy Dindal

Company Contact Roger Jenkins

Telephone (615) 576-8594

Project Designation/Sampling Locations 241-BY-111 (VSS)

Collection Date

Ice Chest No.

Field Logbook No.

Bill of Lading/Airbill No.

Offsite Property No.

Method of Shipment Federal Express

Shipped to Rick Mahon

Possible Sample Hazards/Remarks NONE

Sample Identification

S4083-A10.610	TST# 610 ✓	ORNL Triple Sorbent Trap
S4083-A11.644	TST# 644 ✓	ORNL Triple Sorbent Trap
S4083-A13.645	TST# 645 ✓	ORNL Triple Sorbent Trap
S4083-A14.648	TST# 648 ✓	ORNL Triple Sorbent Trap
S4083-A15.656	TST# 656 ✓	ORNL Trip Blank
S4083-A12.663	TST# 663 ✓	ORNL Field Blank

[] Field Transfer of Custody

Chain of Possession

(Sign and Print Names)

Relinquished By	Date	Time	Received By	Date	Time
AB Dindal / <i>AB Dindal</i>	11/3/94	10:46AM	<i>AP Brubaker</i>	11-4-94	1400
<i>AP Brubaker</i>	11-16-94	0730	<i>T-B Latta / Rick Mahon</i>	11-16-94	0730
<i>T-B Latta / Rick Mahon</i>	11-25-94	1100	<i>AB Dindal / Rick Mahon</i>	11/29/94	1722

Did not include

max/min thermometer

JBS 11/29/94

Temperature (°C)	Min	Max
When shipped (ORNL)	0	12
When received (WHC)	2	6
When shipped (WHC)		
When received (ORNL)		9.6

Westinghouse
Hanford Company

CHAIN OF CUSTODY

WHC 008077

Custody Form Initiator J. A. Edwards-PNL

Telephone (509) 373-0141

Page 85-3009 / FAX 376-0418

Company Contact R. D. Mahon-WHC

Telephone (509) 373-1101

Page 85-9656 / FAX 373-3193

Project Designation/Sampling Locations 200 East Tank Farm
241-BY-111 Tank Vapor Sample SAF S4083
(VSS Truck)

Collection date 11-__-94

Preparation date 11-0 7-94

Ice Chest No.

Field Logbook No. WHC-____-____-____

Bill of Lading/Airbill No. N/A

Offsite Property No. N/A

Method of Shipment Government Truck

Shipped to PNL

Sample Job # _____

Possible Sample Hazards/Remarks Unknown at time of sampling

Sample Identification

S4083 - A16 .91V [^]	NH ₃ /NO _x /H ₂ O (Trap # 1) Line # 9
S4083 - A17 .92V [^]	NH ₃ /NO _x /H ₂ O (Trap # 2) Line #10
S4083 - A18 .93V [^]	NH ₃ /NO _x /H ₂ O (Trap # 3) Line # 8
S4083 - A19 .94V [^]	NH ₃ /NO _x /H ₂ O (Trap # 4) Line #10
S4083 - A20 .95V [^]	NH ₃ /NO _x /H ₂ O (Trap # 5) Line # 9
S4083 - A21 .96V [^]	NH ₃ /NO _x /H ₂ O (Trap # 6) Line #10
S4083 - A22 .97V [^]	NH ₃ /NO _x /H ₂ O a-b-c (Trap Trip Blank# 1)
S4083 - A23 .98V [^]	NH ₃ /NO _x /H ₂ O a-b-c (Trap Trip Blank# 2)
S4083 - A24 .99V [^]	NH ₃ /NO _x /H ₂ O a-b-c (Trap Trip Blank# 3)

[] Field Transfer of Custody			[] Chain of Possession (Sign and Print Names)		
Relinquished By	Date	Time	Received By	Date	Time
Gary Dennis <i>J. A. Edwards</i>	11-09-94	0800	J. A. Edwards <i>J. A. Edwards</i>	11-09-94	0800
J. A. Edwards <i>J. A. Edwards</i>	11-18-94	1400	T. R. Mahon / T. A. Edwards	11-18-94	1400
Tim B. Uhlert / T. R. Mahon	11-28-94	1200	J. A. Edwards / J. A. Edwards	11-28-94	1200

Final Sample Disposition

Comments:

	PNL (only) Checklist	Pick-up / Delivery	Comments:
0	Media labeled and checked?	☒ Y / N	
0	Letter of Instruction?	☒ Y / N	
0	Media in good condition?	☒ Y / N	
0	COC info/signatures complete?	☒ Y / N	
0	Sorbents shipped on ice?	☒ Y / N	
0	Rad release stickers on samples?	☒ Y / N	
0	Activity report from 222S?	☒ Y / N	
0	COC copy for LRB, RIDS filed?	☒ Y / N	
0	COC copy for sorbent follow-on?	☒ Y / N	
	POC ☒	POC ☒	

(Revised 10/17/94 PNL)

Westinghouse Hanford Company	CHAIN OF CUSTODY		008521
Custody Form Initiator <u>T.B. Utecht</u>			
Company Contact <u>R.D. Mahan</u>		Telephone <u>373-1101</u>	
Project Designation/Sampling Locations <u>241-BY Tank 111</u>		Collection Date <u>11-16-94</u>	
Ice Chest No. <u>N/A</u>		Field Logbook No. <u>N/A</u>	
Bill of Lading/Airbill No. <u>N/A</u>		Offsite Property No. <u>N/A</u>	
Method of Shipment <u>Federal Express</u>			
Shipped to <u>O&I</u>			
Possible Sample Hazards/Remarks <u>NONE</u>			
Sample Identification			
<u>54083 - A04. H205</u>			
<u>54083 - A05. H232</u>			
<u>54083 - A06. H233</u>			
☒ Field Transfer of Custody			
CHAIN OF POSSESSION			
(Sign and Print Name)			
Relinquished By	Date	Time	Received By
<u>Tim B. Utecht / E. B. Utecht</u>	<u>11-28-94</u>	<u>1200</u>	<u>JA Edwards / K. Edwards</u>
Final Sample Disposition			
Disposal Method:		Disposed by:	Date/Time:
Comments:			

WHC-SD-WM-RPT-124, Rev. 0

Westinghouse
Hanford Company

CHAIN OF CUSTODY

WHC 008076

Custody Form Initiator

J. A. Edwards-PNL

Telephone (509) 373-0141

Page 85-3009 / FAX 376-0418

Company Contact

R. D. Mahon-WHC

Telephone (509) 373-1101

Page 85-9656 / FAX 373-3193

Project Designation/Sampling Locations 200 East Tank Farm

241-BY-111 Tank

Vapor Sample SAF S4083

(VSS Truck)

Collection date 11-12-94 (11-28-94)

Preparation date 11-10-94

Ice Chest No.

Field Logbook No. WHC-_____-

Bill of Lading/Airbill No.

N/A

Offsite Property No. N/A

Method of Shipment

Government Truck

Shipped to

PNL

Sample Job # _____

Possible Sample Hazards/Remarks Unknown at time of sampling

Sample Identification

S4083-A01.223~

S4083-A02.224~

Ambient Air #1 Upwind @ VSS (PNL)

Ambient Air #2 Through VSS (PNL)

S4083-A07.247~

S4083-A08.248~

S4083-A09.249~

Sample #6 (PNL)

Sample #7 (PNL)

Sample #8 (PNL)

[] Field Transfer of Custody			[] Chain of Possession			(Sign and Print Names)	
Relinquished By	Date	Time	Received By	Date	Time		
J. A. Edwards	11-10-94	1400	JP Bunch	11-10-94	1400		
JP Bunch	11-11-94	0730	TRIMBLE/T. R. L. M.	11-11-94	0730		
T. Bunch/T. R. L. M.	11-25-94	1200	J. A. Edwards/STA Edwards	11-28-94	1200		

Final Sample Disposition

Comments:

	PNL (only) Checklist	Pick-up / Delivery	Comments:
0	Media labeled and checked?	(Y) N	
0	Letter of instruction?	(Y) N	
0	Media in good condition?	(Y) N / (Y) N	
0	COC info/signatures complete?	(Y) N / (Y) N	
0	Sorbents shipped on ice?	Y N / Y N	N/A
0	Rad release stickers on samples?	/ (Y) N	
0	Activity report from 222S?	/ (Y) N	
0	COC copy for LRB, RIDS filed?	/ (Y) N	
0	COC copy for sorbent follow-up?	/ Y N	
	FOC (Y)	POC (Y)	

(Revised 10/17/94 PNL)