DOE/ER/61195-5

**DOSIMETRY AND QUANTITATIVE RADIONUCLIDE IMAGING
IN RADIOIMMUNOTHERAPY**

FINAL REPORT

for period July 15, 1992 to July 14, 1996

Peter K. Leichner, Ph.D.
Professor of Radiation Oncology
Department of Radiation Oncology
University of Nebraska Medical Center
Omaha, Nebraska 68198-1050

September 1996

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Prepared for

**THE U.S. DEPARTMENT OF ENERGY
AGREEMENT NO. DE-FG02-91ER61195**

MASTER

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, make any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

**Portions of this document may be illegible
in electronic image products. Images are
produced from the best available original
document.**

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the Department of Energy, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product or process disclosed or represents that its use would not infringe privately-owned rights.

ABSTRACT

This final report under agreement No. DE-FG02-91ER61195 summarizes the overall progress in our research on the dosimetry and quantitative radionuclide imaging in radioimmunotherapy in the following areas:

1. Development of quantitative SPECT for high-energy photons (e.g. ^{87}Y , ^{19}F) and stability testing of ^{87}Y -labeled antibodies in the nude mouse model. This work has demonstrated the feasibility of using ^{87}Y -labeled antibodies for imaging prior to therapy with ^{90}Y -labeled antibodies.
2. Development of a unified approach to photon and beta particle dosimetry. The point-source kernel in this approach can be used for photons and beta particles for uniform and non-uniform activity distributions.
3. Quantitative SPECT for nonuniform attenuation. In a phantom study with non-uniform attenuation, differences between measured and computed activity concentrations were less than 6%.
4. Development of patient-specific dosimetry in radioimmunotherapy. Patient-specific dosimetry is based entirely on patients' quantitative SPECT studies and whole-body imaging and obviates the need for using "standard organs" in radiation dosimetry.

SUMMARY OF OVERALL PROGRESS

The results of our research have been published or have been accepted for publication in the open literature. Consistent with our previous reports, we therefore provide only brief summaries that describe the salient features of the approaches employed and the results obtained.

1. The use of ^{87}Y in radioimmunotherapy

In several clinical trials, ^{111}In -labeled antibodies were used for imaging and dosimetry prior to therapy with ^{90}Y -labeled antibodies. It is unlikely that the pharmacokinetics of ^{111}In - and ^{90}Y -labeled antibodies were exactly the same. Therefore, radiation dosimetry may have been compromised. To overcome this problem, we developed quantitative SPECT for ^{87}Y because ^{87}Y and ^{90}Y have the same chemical properties and chelation chemistry. Additionally, the in-vivo stability of ^{87}Y radioimmunoconjugates was determined by conjugating monoclonal antiferritin antibody to the chelator ITCB-DTPA and chelating with ^{87}Y . The results of these studies (see Publications 4 and 5) demonstrated that in a phantom study, ^{87}Y -containing fillable "organs" were imaged very well and that ^{87}Y activities in these organs were computed with a difference of less than 10% as compared to actual measurement. In tumor-bearing nude mice, tumor targetting and normal-organ distribution of ^{87}Y - and ^{90}Y -labeled antibodies were the same. The imaging results and animal data suggest that ^{87}Y -labeled antibodies could be used to reliably predict the pharmacokinetics of ^{90}Y -labeled antibodies and, hence, improve dosimetry in radioimmunotherapy.

In a subsequent investigation it was shown that the collimators used for imaging ^{87}Y photons also have potential application in ^{18}F (positron) imaging. A complete characterization of the collimators used is provided in Publication No. 21 (reprint enclosed).

2. A unified approach to photon and beta particle dosimetry

A complete description of this research is provided in Publication No. 16; a reprint is enclosed.

The objective of this work was to develop a unified and practical method for photon and beta particle dosimetry. This was achieved by developing a point-source kernel that is equally valid for photons and beta particles. Explicit expressions were derived for the absorbed fraction within and outside of spheres containing a uniform distribution of activity. An important feature of the derivations is that absorbed-dose calculations can be

made analytically on the macroscopic, cellular and subcellular levels. As outlined below, the methodology employed and results obtained were subsequently utilized in patient-specific tumor and normal-organ dosimetry.

3. Quantitative SPECT for nonuniform attenuation

An important goal of our research was to develop and test quantitative SPECT for nonuniform attenuation with the circular harmonic transform (CHT) reconstruction algorithm developed by our group. Mathematically, this was achieved by developing a three-dimensional collimator filter model and employing the frequency-distance principle in CHT reconstructions.

Nonuniform attenuation is most pronounced in the thorax. Therefore, a phantom was constructed to simulate the different attenuations found in this part of the body. A detailed description of this phantom was provided in Progress Report DOE/ER/61195-3, March 1994. As summarized in that report, the maximum difference between actual and computed radionuclide concentrations was approximately 6%.

4. Development of patient-specific dosimetry

The development of patient-specific dosimetry was the ultimate goal of our research in the dosimetry of administered radionuclides in general, and specifically, radiolabeled antibodies. A detailed description of the methodology used and results obtained is provided in Publications 24 and 26.

Briefly, this approach to patient dosimetry is based on whole-body imaging and quantitative SPECT to generate time-activity curves for those tumors and normal organs which demonstrate significant uptake of radiolabeled antibodies. Additionally, tumor and normal organ volumes and the corresponding activity concentrations are determined from transverse SPECT slices. The absorbed doses corresponding to the activity concentrations are calculated using a three-dimensional discrete Fourier transform (3D-DFT). The 3D-DFT convolution method was validated in mathematical and physical phantoms by comparing it to Monte Carlo transport calculations which were carried out using the EGS4 system code. The root mean square error between the two methods was less than 0.1%. It was concluded that the 3D-DFT method which is computationally much more efficient than Monte Carlo transport calculations is a precise tool for carrying out absorbed-dose calculations in clinical studies.

PUBLICATIONS ISSUED DURING THE TOTAL AGREEMENT PERIOD

1. Rahmouni A, Yang A, Tempany CMC, Frenkel T, Epstein J, Walsh P, Leichner PK, Ricci C, Zerhouni E. Accuracy of in-vivo assessment of prostatic volume by MRI and ultrasonography. J Comput Assist Tomogr 16:935-940, 1992.
2. Hawkins WG, Yang N-C, Leichner PK. Theorems of the number of zeros of the radial modulators of the 2D exponential radon transform. In: Lecture notes in mathematics, Springer Verlag, 1497:194-200, 1992.
3. Hawkins WG, Links JM, Lam WC, Leichner PK. FFT interpolation. J Nucl Med 33:1007, 1992.
4. Leichner PK, Harvey EB, Holdeman KP, Harrison KIA, Hawkins WG, Kelly RF, Atcher RW, Lexa R, Valentino F, Dalrymple GV. SPECT imaging of Y-87 (485-keV) photons. Antibody Immunoconj Radiophar 5:355, 1992.
5. Quadri SM, Lai J, Ansari MH, Atcher RW, Leichner PK. Evaluation in a nude mouse model of Yttrium-87 labeled monoclonal antiferritin antibody as a potential radioimmunoimaging agent. Antibody Immunoconj Radiophar 5:350, 1992.
6. Leichner PK, Kwok CS. Tumor dosimetry in radioimmunotherapy: Methods of calculation for beta particles. Med Phys 20:527-534, 1993.
7. Leichner PK, Koral KF, Jaszczak RJ, Green AJ, Chen GTY, Roeske JC. An overview of imaging techniques and physical aspects of treatment planning in radioimmunotherapy. Med Phys 20:569-577, 1993.
8. Quadri SM, Vriesendorp HM, Leichner PK, Williams JR. Evaluation of Indium-111 and Yttrium-90-labeled linker-immunoconjugates in nude mice and dogs. J Nucl Med 34:938-945, 1993.
9. Quadri SM, Shao Y, Blum JA, Leichner PK, Williams JR, Vriesendorp HM. Preclinical evaluation of ¹¹¹In- and ⁹⁰Y-labeled immunoconjugate (GYK-DTPA) in beagle dogs. Nucl Med Biol 20:559-570, 1993.
10. Tempero M, Colcher D, Dalrymple G, Harrison K, Shantaram J, Quadri S, Linder J, Augustine S, Reed E, Leichner P. Cancer therapy using radioimmunoconjugates: Implications for breast cancer. Ann New York Acad Sci 698:406-417, 1993.
11. Tempero M, Colcher D, Dalrymple G, Harrison K, Augustine S, Schlom J, Lindner J, Leichner PK. High dose therapy with ¹³¹I conjugated monoclonal antibody CC49: A phase I trial. Antib Immunoconj Radiophar 6:90, 1993.

12. Harrison KA, Dalrymple GV, Leichner PK, Tempero MA, Colcher DM, Augustine SC, Bierman PJ, Joshi SS, Holdeman KP, Quadri SM, Schneiderman MH, Jacobson DA. Clinical radiobiology of radiolabeled antibody therapy. Proceedings 13th Annual Meeting of the Radiation Research Society, p.117, 1993.
13. Schneiderman MH, Schneiderman GS, Dalrymple GV, Leichner PK. ^{123}I UdR incorporation into DNA. Proceedings 13th Annual Meeting of the Radiation Research Society, p.154, 1993.
14. Dalrymple GV, Leichner PK, Harrison KA, Cox AB, Hardy KA, Salmon YL, Mitchell JC. Induction of high-grade astrocytoma (HGA) by protons: Molecular mechanisms and RBE considerations. Adv Space Res 14:267-270, 1994.
15. Hawkins WG, Links JM, Leichner PK. FFT interpolation for arbitrary factors: Application to cross-sectional imaging. Proceedings 14th world congress on computational and applied mathematics 2:274-277, 1994.
16. Leichner PK. A unified approach to photon and beta particle dosimetry. J Nucl Med 35:1721-1729, 1994.
17. Harrison K, Bierman P, Leichner P, et al. ^{90}Y antiferritin for end-stage Hodgkin's disease. Antibody Immunoconj Radiophar 7:92, 1994.
18. Dalrymple GV, Schneiderman MH, Baranowska-Kortylewicz J, Leichner PK, et al. Auger electron emitters for cancer imaging prior to therapy. Proceedings. 14th annual meeting of the Radiation Research Society, p.251, 1994.
19. Harrison K, Dalrymple GV, Baranowska-Kortylewicz J, Schneiderman MH, Holdeman K, Leichner PK, Augustine SC, Jacobson D. Bladder cancer [^{123}I]UdR imaging in preparation for [^{125}I]UdR therapy. J Nucl Med 35:144P, 1994.
20. Yang N-C, Hawkins WG, Leichner PK. Comparisons of spatial domain filters and frequency domain filters on CT images of primary liver cancers. Med Phys 21:904, 1994.
21. Leichner PK, Morgan HT, Holdeman KP, Harrison KA, Valentino F, Lexa R, Hawkins WG, Dalrymple GV. SPECT imaging of fluorine-18. J Nucl Med 36:1472-1475, 1995.
22. King DJ, Antoniow P, Owens RJ, et al. Preparation and preclinical evaluation of humanised A33 immunoconjugates for radioimmunotherapy. Brit J Cancer 72:1364-1372, 1995.
23. Harrison KA, Dalrymple GV, Baranowska-Kortylewicz J, et al. Radiolabeled

iododeoxyuridine: Safety evaluation. J Nucl Med 37 (Suppl):13S-16S, 1996.

24. Leichner PK, Akabani G, Colcher D, Harrison KA, Hawkins WG, Eckblade M, Baranowska-Kortylewicz J, et al. Patient-specific dosimetry of ^{111}In - / ^{90}Y -labeled monoclonal antibody CC49. In press, J Nucl Med, 1996.
25. Leichner PK. Radiation dosimetry of radiolabeled monoclonal antibodies: Practical considerations. In: Nuclear Medicine, Vol. 1, Henkin RE, et al., eds., Mosby, New York, 558-562, 1996.
26. Akabani G, Hawkins WG, Eckblade M, Leichner PK. Patient-specific dosimetry using quantitative SPECT and 3D-FT convolution. In press, J Nucl Med, 1996.
27. Akabani G, Hawkins WG, Leichner PK. Small-scale alpha dosimetry of red bone marrow using histological images. J Nucl Med 37:231P, 1996.

APPENDIX
DOE/ER/61195-5

Reprints of two articles:

1. Leichner PK. A Unified Approach to Photon and Beta Particle Dosimetry.
J Nucl Med 34:1721-1729, 1994.
2. Leichner PK, Morgan Ht, Holdeman KP, et al. SPECT Imaging of Fluorine-18.
J Nucl Med 36:1472-1475, 1995.

*Pulled
for separate
processing*