LA-UR- 96 - 4901

* Approved for public release;
distribution is unlimited.

Title:

The Laser Safety Program at Los Alamos
National Laboratory

CONF - 970354 --1

Author(s):

Ron Hyer

RECEIVED

JAN 21 1997

OSTI

Submitted to:

Laser Institute of America

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Los Alamos
NATIONAL LABORATORY

Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by the University of California for the U.S. Department of Energy under contract W-7405-ENG-36. By acceptance of this article, the publisher recognizes that the U.S. Government retains a nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. The Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

DISCLAIMER

**Portions of this document may be illegible
in electronic image products. Images are
produced from the best available original
document.**

12/14/98 09:55 2000007004 ESH-5 Op. Safety 444 NAME CLASSIFICA 0004

Abstract for a paper to be presented at the 1997 International Laser Safety Conference, March 17-20, 1997, Orlando, Florida. Submitted to Laser Institute of America, 12424 Research Parkway, Suite 125, Orlando, FL 32826

The Laser Safety Program at Los Alamos National Laboratory

Ron Hyer, Laser Safety Officer
Laser Safety Officer
Los Alamos National Laboratory
Los Alamos, New Mexico
United States

ABSTRACT

The Laser Safety Program at the Los Alamos National Laboratory was formalized in April, 1991, with the publication of a document, "Lasers," modeled on the ANSI Z136.1 standard. This program has received such wide acceptance by the laser community and line managers that the original Laser Safety Program document has become a Laboratory standard on lasers.

As a benchmark of the success of this program is that the Laboratory has experienced no disabling eye injuries because of laser operations since July, 1990, to be compared with a disabling laser eye injury that used to average one every eighteen months prior to the time the formal program was established.

The Laboratory Laser Safety Program and program elements will be presented and discussed.

12/24/88 08:33 0000000004 ESH-5 Op. Safety 333 LANL CLASSIFICA 005

The Laser Safety Program at Los Alamos National Laboratory

Ron Hyer, Los Alamos National Laboratory, Los Alamos, NM 87545

SUMMARY

Work with lasers and laser systems must be performed with minimal exposure to laser hazards. Los Alamos National Laboratory (LANL) organizations and personnel are required to comply with Department of Energy (DOE) Order 5480.4, "Environmental Protection, Safety, and Health Protection Standards," that requires laser activities be performed in compliance with American National Standards Institute (ANSI) Z136.1, "American National Standard for the Safe Use of Lasers." Laboratory personnel must also comply with the provisions of the written laser safety program, "Lasers," which supplements the ANSI Z136.1 standard. The implementation of these requirements in a research and development environment is the subject of the following paper.

Responsibility for the LANL Laser Safety Program

The responsibility for the LANL laser safety program is shared among the laser safety officers (LSOs), the line manager, the facility manager, and the laser users.

The laser safety officer (LSO) is required to:

- ◆ Perform the duties specified in ANSI Z136.1 and the LANL written laser safety program, "Lasers."
- ◆ Ensure that laser operations within his or her jurisdiction comply with the requirements of ANSI Z136.1.
- Monitor and enforce the control of laser hazards, and
- ◆ Ensure that all documentation required by the written program for laser operations is accurate and submitted in a timely manner.

The line manager and facility manager are required to:

- ◆ Ensure that the cognizant LSO is informed of all laser operations within his or her authority;
- ◆ Ensure that laser workers are given adequate training for safe laser operation and that this training is documented according to Laboratory standards;

- ◆ Ensure that all documentation, evaluations, reviews, and approvals required by the written program are verified before the operation of any laser under his or her authority;
- Identify and select personnel who require training for safe laser operation; and
- Allocate resources for identified training.

Training of Laser Personnel

Only trained and authorized personnel are permitted to operate class 3b and class 4 lasers and laser systems. Line managers and facility managers must establish operational requirements for their laser operations.

To meet these requirements, training courses are offered by the Environment, Safety, and Health (ES&H) training group. Laser safety orientation is a course required for all laser and incidental personnel who work directly or around class 3b or class 4 lasers. Before beginning work with class 3b or class 4 lasers, laser personnel must complete the required training course or receive an interim laser safety briefing from the cognizant qualified laser operator and the cognizant LSO on the hazards and controls of the specific laser in use and the contents of the Laboratory written laser safety program. Personnel must complete the required training course within 90 days after receiving the interim laser safety briefing. After personnel have completed the laser safety orientation course, they must complete a refresher course every three years.

LSOs must complete the laser safety officer course before assuming authority. Retraining is required every three years and can be accomplished by attending a laser safety officer training course or by other means of recertification approved by the Laboratory Laser Safety Officer.

The Laboratory required laser safety training is viewed as the minimum requirement. Additional laser education and training may be warranted in some situations, based upon the particular situation and previous laser training, education, and experience of the personnel involved. Where applicable, training should include electrical safety and cardiopulmonary resuscitation (CPR).

Laser personnel and incidental personnel working in class 3b and class 4 laser operation areas must have initial eye examinations as soon as possible after receiving their work assignments. Eye examinations are not required for personnel working in class 1, class 2, or class 3a operation areas. An examination is required following any suspected eye exposure to lasers. No periodic re-examinations are required.

Hazard Evaluation

Evaluation of laser hazards must be performed in accordance with ANSI Z136.1 and the LANL written laser safety program, "Lasers." This is a joint effort between the laser user and the LSO. For lasers that require an SOP/SWP for operation, an LSO must evaluate the hazard and

report the results on Form 1552, "Laser Registration and Hazard Evaluation Report." The laser user must provide the LSO with accurate information concerning the performance of the laser and mode of operation.

An LSO is the only person authorized to classify or reclassify lasers at the Laboratory. Lasers and laser systems must be classified and labeled according to the requirements of ANSI Z136.1. Manufacturer's classification and labeling of a laser in conformance with 21CFR1040, "Performance Standards for Light-Emitting Products," may be used to satisfy this requirement, but an LSO must classify and label lasers that have been built at the Laboratory.

An LSO must reclassify and relabel a laser if the output power, energy, or other operating characteristics have changed because of modification, enhancement, or degradation of the laser, causing the level of hazard to change.

An SOP/SWP is required for all operations that use any class 3b or class 4 laser, except those continuous-wave (CW) lasers with a wavelength between 400 nanometers and 750 nanometers and an accessible radiant power less than 50 milliwatts.

Alignment procedures must be included in all laser SOPs for operations that require the alignment of laser optical systems (such as mirrors, lenses, detectors, and optical cavities). The procedure must include steps that minimize the potential for exposure to hazardous laser radiation during an alignment operation.

The approved SOPs must be prominently posted in the laser operation area. All personnel involved in operating and maintaining the laser or laser system must document that they have read and understood the SOP.

If the laser hazard analysis indicates that the maximum level of accessible radiation is that of a class 3b or class 4 laser, then the cognizant LSO must establish a laser controlled area and require laser control measures in accordance with ANSI Z136.1.

A temporary laser controlled area must be established if an embedded class 3b or class 4 laser is activated when the enclosure is open for maintenance or service.

Access to laser operation areas is controlled by access-control interlocks, access-control light panels, warning signs and procedures.

Access control interlocks on doors or other means of entrance into a class 3b or class 4 laser controlled area are required if one or more of the following conditions exist:

- ◆ Engineered barriers that prevent exposures have not or cannot be implemented;
- ◆ Hazardous laser radiation, including hazardous diffuse reflections, may be present at the doorway;

- Eye protection is required as a primary hazard control when lasers are on; or
- Any other hazard is present that the LSO determines would be significantly reduced by implementing access-control interlocks.

Appropriate control measures reduce the possibility of exposing personnel to hazards associated with lasers and laser systems during operations, maintenance, and service.

Engineering control measures are preferred when selecting control measures for limiting access to laser radiation and associated hazards.

Laser protective eyewear is prescribed during all class 3b and class 4 laser operations when other control measures do not prevent access to direct laser beams, specular reflections, or hazardous diffuse reflections. As part of the LSO evaluation of the laser beam hazard, the need for eye protection is determined and required for operations.

In addition to direct hazards to eyes and skin from laser beam exposure, other hazards associated with the use of lasers must be controlled. Some examples of non-beam hazards are the following:

- Electrical;
- Laser generated air contaminants (LGACs);
- Collateral radiation;
- Fire;
- Noise;
- Confined spaces;
- Ergonomic problems; and
- Cryogenic liquids.

Laser dyes and dye solvents vary considerably in toxicity and carcinogenicity and are treated as hazardous chemicals. Special care is taken when handling these materials to reduce the possibility of personnel exposure by following proper industrial hygiene practices to safely handle these materials.

Results and Conclusion

One of the most useful enhancements to the overall laser safety program has been in the area of standardized laser safety training, using a computer based training course to teach and test the knowledge of laser users. Since this course is computer based, no instructor is required and the student can schedule and receive the training more or less at his or her convenience. This training course received has been effective and well received by students who rank it highly on their course evaluations.

An independent assessment of the LANL laser safety program was completed in May, 1996. The criteria for this assessment were based upon the LANL written laser safety program

and consisted of interviews with group, division, and facility laser safety officers, laser operators and supervisors, and observations of operational laser systems. The results pointed out potential improvements to the current laser safety program that were mainly administrative having to do with roles and responsibilities of group, division and facility management of internal laser safety issues, accountability of laser systems as to type, ownership, status and location, and medical surveillance issues.

In 1990 the goal of the LANL laser safety program was to eliminate disabling eye injuries that were occurring at the alarming interval of one every 18 months and serious injuries or fatalities as the result of working with lasers. The ANSI standard, Z136.1-1986, was used as a basis for the first LANL written laser safety program, AR 5-2, "Lasers," and TB501, "Laser Safety," that later became separate sections in the non-ionizing radiation part of the LANL Environment, Safety and Health Manual. Recently, these two documents have been combined into a draft Laboratory Standard, LS106-14.0, "Lasers," that is in the final approval process. As already mentioned in the beginning, the last serious accident involving a laser was a disabling eye injury that occurred at the LANL in July 1990. The improved safety record is attributed to management commitment and increased awareness of laser hazards that has resulted to improved administrative and engineering controls, better engineered products with built in safety features and safety training for users.

Autobiography

Ron Hyer began his career at Lawrence Livermore National Laboratory, Livermore, CA, where he worked in optical diagnostics of weapons tests. With the advent of the laser he developed an interest in laser science and transferred to Los Alamos National Laboratory, Los Alamos, NM, where he worked on the development of a large Nd: Glass laser system as part of the laser fusion program. He became a safety engineer in 1989 and now serves as the Laboratory laser safety officer.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.