DOE/NV/11665-T1

**SOIL CONTAMINATION STUDIES
DE-FG08-94NV11665**

June 1997

University of Nevada, Las Vegas
Department of Health Physics

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

MASTER

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

1. INTRODUCTION

The University of Nevada, Las Vegas (UNLV) Department of Health Physics was awarded grant DE-FG08-94NV11665 entitled "Soil Contamination Studies" on October 14, 1994. The objective of the project was to develop a quick screening method that accurately identifies and quantifies the amount of alpha-emitting radionuclides in infinitely-thick soil samples using a Frisch grid ionization chamber. An additional objective of the work was to provide the U.S. Department of Energy, Nevada Operations Office (DOE/NV) and its contractors with information on the theoretical and actual measured results of atmospheric testing contamination of soil and water at the Nevada Test Site (NTS) through a comprehensive search of existing literature.

2. Determination of Alpha-Emitting Radionuclides in Infinitely-Thick Soil Samples

Soil sample preparation techniques for alpha spectroscopy are often quite tedious and labor intensive. Because of the extensive use of chemicals, and subsequently, required use of fume hoods, current methods are fine for laboratory settings but unsuitable for timely field assay analyses. The Frisch grid gas ionization chamber has shown promise for near real-time analysis of in-field environmental soil samples, with minimal space and cost requirements.

There are several advantages to using a Frisch grid system for alpha spectrometry work. It has a large interior counting chamber that can accommodate several sample geometries. The chamber has a full 2 pi geometry allowing the counting gas to completely circulate around the sample resulting in efficiencies approaching 50%, which lowers detection limits and minimizes required counting times. Its resolution is comparable to that of conventional solid state alpha detectors. An electroplated alpha standard (^{239}Pu , ^{241}Am , and ^{244}Cm) was counted both on a Frisch grid detector and a state-of-the-art solid state detector to compare the efficiency and resolution of these two types of systems. Results shown in Figures 1 and 2 indicated that the Frisch grid used in this study provided efficiencies of 40-45% versus an efficiency of 24% for a typical solid-state silicon detector. The Frisch grid gave a resolution of 3.31 keV/channel for the ^{239}Pu peak versus 2.7 keV/channel for a solid state silicon detector.

Because the Frisch grid can accommodate odd sample geometries, it can be used for rapid quantification of environmental samples, specifically soil samples, when a quick determination of alpha contamination is needed. However, determining the total alpha activity of individual radionuclides in infinitely-thick soil samples with a Frisch grid ionization chamber presents a number of problems. First, the introduction of loose, uncovered alpha-emitting material in the sensitive volume of the detector may

Figure 1. Frisch grid spectrum obtained with 5 nCi ^{239}Pu , ^{241}Am , and ^{244}Cm electroplated standard. Efficiency for ^{239}Pu = 40%, resolution = 3.31 keV/channel. 300-sec count.

Figure 2. Solid-state silicon detector spectrum obtained with 5 nCi ^{239}Pu , ^{241}Am , and ^{244}Cm electroplated standard. Efficiency for ^{239}Pu = 24%, resolution = 2.7 keV/channel. 300-sec count.

result in the contamination of the detector's sensitive volume. The extent of detector contamination is further exacerbated when contaminated soil particles are aerosolized during purge/fill cycles of the P-10 counting gas and adhere on counting chamber surfaces. Increases in background levels decrease the resolution of the detection system. In addition, decontamination of a Frisch grid detector chamber is very time consuming and requires great care to ensure delicate features of the detector such as the grid are not damaged.

Full alpha energy peaks in the spectra taken of infinitely-thick soil samples in a Frisch grid detector are often convoluted due to tailing effects and superposition of monoenergetic peaks as shown in Figure 3. The low energy tailing effect is caused by the partial energy loss of alpha particles caused by interaction with the soil particles. This prevents full energy deposition of the alpha particle in the detector sensitive volume and leads to a drastic reduction in resolution. To attain sufficient resolution between pulse height distributions, soil samples must be counted for extended periods of time, or very small sample sizes must be used. If precautions are taken and full energy peaks are discernable in the spectrum, commercially-available software programs that deconvolute spectra can be utilized to accurately determine net counts in each peak.

In this study, a number of soil sample preparation techniques for counting on a Frisch grid ionization chamber were investigated. Planchettes containing arid soil samples collected from the Double-Tracks Site at the Nevada Test Site were treated with various materials prior to counting to determine the optimal means for preventing aerosolization of the soil in the sensitive volume of the detector and preventing/minimizing tailing effects in spectra during analysis. The soil was assayed in the field and was estimated to have a specific activity of ^{239}Pu of approximately 29 nCi/g.

The Frisch grid used was a model Adiago 114 manufactured by Numelec-Instruments. The detector was constructed to run through 2 purge/fill cycles per sample which lasted for 175 seconds. Each sample was counted for 20 minutes. A 5-minute background was taken after each sample was counted to determine if any contaminated soil had aerosolized during analysis.

The sample preparation methodologies investigated as part of this study include:

1. UNLV radiochemical procedure (baseline)
2. Dispersing soil in methanol
3. Dispersing soil in a methanol/nitric acid solution
4. 3M adhesive
5. Double-sided tape

Figure 3. Low energy tailing effects typically seen in spectrum by a Frisch grid spectrum upon analysis of thick, ^{239}Pu -contaminated soil sample. 1200-sec count.

Methanol Methodology

This methodology was adopted from Sill¹ and was demonstrated during preliminary cleanup efforts at the Radioactive Waste Management Complex at the Idaho National Environmental and Engineering Laboratory. The modified procedure involves adding 2 mL of methanol to a known mass of contaminated soil on a 2-inch diameter aluminum planchet and then heating the sample to dryness on a hotplate. The methanol acts to disperse the soil on the planchet to reduce soil clumping, and subsequently, alpha particle absorption. The 5-minute backgrounds obtained after each sample analysis indicated that some of the contaminated soil had aerosolized during the analysis. The methanol methodology proved to be inadequate for binding the contaminated soil to the planchet during the purge/fill cycles. Sill successfully fixed soil on planchets with his method by grinding the soil after the methanol application. This approach was not used in this study to prevent the generation of airborne, respirable ²³⁹Pu-contaminated soil particulates.

Methanol/Nitric Acid Methodology

This methodology was adopted from Lally² and involves adding methanol and 1 M nitric acid to a known mass of contaminated soil on a 2-inch diameter aluminum planchet and then heating the sample to dryness on a hotplate. Both the methanol and nitric acid act to disperse the soil on the planchet to reduce soil clumping. The nitric also provides some initial leaching of the soil surfaces to further disperse the contamination. The 5-minute backgrounds obtained after each sample analysis indicated that there was no contamination of the Frisch grid chamber. The actual procedure used is outlined below:

1. Clean 2" diameter aluminum planchet with methanol and allow to dry.
2. Weigh dry planchet and record weight.
3. Pass soil sample through 50-mesh screen (250 μ m) and set aside.
4. Place planchet on scale and transfer no more than 0.25 g of sample and record weight.
5. Add 2 mL of methanol to soil/planchet and swirl gently to disperse sample and place on WARM hotplate. If the hotplate is too warm, the methanol will splatter.
6. Add 1 mL of 1 M nitric acid and swirl planchet to disperse sample.

Note: Additional nitric acid can be added if necessary. Volumes can be adjusted as needed depending on the area and

volume of the planchet. It is not recommended adding more liquid while planchet is in direct contact with hotplate because splattering can occur.

7. Allow planchet to dry completely. This can be checked by re-weighing after drying.

Figures 4-6 are examples of spectra obtained by counting Double Tracks soil on the Frisch grid for 20 minutes. Figures 4 and 5 show a relatively well defined ^{239}Pu peak at 5.15 MeV with some tailing effects at lower energies. Note also in both cases that there is a fairly well defined ^{241}Am peak at 5.40 MeV. The Double Tracks soil has a high enough specific activity that it takes only ~ 0.01 g of soil to actually see ^{239}Pu activity. This suggests that rapid quantification (i.e., 20 minutes) of plutonium in soils using the methanol/nitric acid method and a Frisch grid detector can only be obtained in high specific activity soils. Another weakness of any Frisch grid analysis of soils, regardless of preparation technique, is that the presence of too many alpha-emitting radionuclides may convolute a spectrum to the point where quantification may be impossible.

Note that the spectrum in Figure 6 shows poor resolution and a large amount of low energy tailing. This most likely is due to a "geometry effect." It appears that in some cases, the ^{239}Pu may be sorbed on the underside of a soil particle causing its alpha emissions to be partially absorbed by the soil. This will be a shortcoming of any of the soil preparation methods and will need to be investigated in the future. Finally, the three methanol/nitric samples soil samples tested showed a relatively large variation in their activities even though they were taken from the same 100-g soil sample (14.6 nCi/g, 65.7 nCi/g, and 9.04 nCi/g, average - 28.6 nCi/g). Soils contaminated by discrete fallout are known for their inherent inhomogeneity and this problem is compounded when the sample sizes are in the range of only ~0.05 g. The small sample size necessary to ensure sufficient soil dispersion is one drawback of quantifying alpha-contaminated soils by the methanol/nitric acid method.

3M Adhesive Methodology

This procedure involves coating a 2-inch diameter aluminum planchet with 3M spray adhesive, adding a known mass of soil to the planchet, and then heating the sample to dryness on a hotplate. The adhesive spray does not provide any dispersion of the soil. The 5-minute backgrounds obtained after each sample analysis indicated that there was no contamination of the Frisch grid chamber. Figures 7-9 are examples of spectra obtained by counting Double Tracks soil on the Frisch grid for 20 minutes. Organic compounds in the spray adhesive appear to offgas and quench the ionization events in the chamber which causes a marked increase in low energy tailing and poor resolution. Note that the offgasing is so severe that there is no discernable ^{239}Pu peak at 5.15 MeV in any of the spectra.

Figure 4. Spectrum obtained from analysis of a 0.0174-g Double Tracks soil sample after sample preparation by methanol/nitric acid methodology. 14.6 nCi/g. Efficiency = 19.2%. 1200-sec count.

Figure 5. Spectrum obtained from analysis of a 0.0138-g Double Tracks soil sample after sample preparation by methanol/nitric acid methodology. 65.7 nCi/g. Efficiency = 19.2%. 1200-sec count.

59 -

50 -

42 -

34 -

25 -

17 -

8 -

0

2000.0

4000.0

Energy (keV)

8000.0

10000.0

12000.0

Figure 6. Spectrum obtained from analysis of a 0.0132-g Double Tracks soil sample after sample preparation by methanol/nitric acid methodology. 9.04 nCi/g. Efficiency = 20.3%. 1200-sec count.

59

50

42

34

25

17

8

0

2000.0

4000.0

6000.0

8000.0

10000.0

12000.0

Energy (keV)

Figure 7. Spectrum obtained from analysis of a 0.0130-g Double Tracks soil sample after sample preparation by 3M adhesive methodology. 1200-sec count.

Figure 8. Spectrum obtained from analysis of a 0.0260-g Double Tracks soil sample after sample preparation by 3M adhesive methodology. 1200-sec count.

Figure 9. Spectrum obtained from analysis of a 0.0334-g Double Tracks soil sample after sample preparation by 3M adhesive methodology. 1200-sec count.

Double-Sided Tape Methodology

This procedure involves placing a 4 cm² section of double-sided tape in the center of a 2-inch diameter aluminum planchet and adding a known mass of soil to the tape. The double-sided tape prevented any dispersion of the soil. The 5-minute backgrounds obtained after each sample analysis indicated that there was no contamination of the Frisch grid chamber. Double-sided tape was investigated because it was thought to offer the same amount of adhesion as the 3M spray, but with no offgasing effects. Figures 10-12 are examples of spectra obtained by counting Double Tracks soil on the Frisch grid for 20 minutes. There is a large amount of low energy tailing and poor resolution associated with each of these spectra. The double-sided tape methodology did not perform well primarily because it was difficult to disperse the soil by hand on the planchet.

Conclusions

Preliminary results indicate that a methanol/nitric acid sample preparation methodology can be used to identify and quantify the amount of plutonium isotopes and other alpha-emitting radionuclides in infinitely-thick arid soil samples using a Frisch grid ionization chamber. Limiting factors include the inhomogeneity of contamination on the soil, sample size, and the presence of a large number of alpha-emitting radionuclides in the soil sample. Despite these limitations, the Frisch grid detector continues to show promise for near real-time analysis of in-field environmental soil samples.

Figure 10. Spectrum obtained from analysis of a 0.0260-g Double Tracks soil sample after sample preparation by double-sided tape methodology. 1200-sec count.

Figure 11. Spectrum obtained from analysis of a 0.0326-g Double Tracks soil sample after sample preparation by double-sided tape methodology. 1200-sec count.

Figure 12. Spectrum obtained from analysis of a 0.0272-g Double Tracks soil sample after sample preparation by double-sided tape methodology. 1200-sec count.

3. Literature Search

A literature search on theoretical and actual measured results of atmospheric testing contamination of soil and water at the NTS was completed in July 1995. Literature was obtained from the DOE/NV reference library, Reynolds Electrical and Engineering Coordination and Information Center, and the University and Community College System of Nevada library system. All reports and documents secured from the search were immediately delivered to Reynolds Electrical and Engineering Company (currently Bechtel Nevada) waste management personnel in support of their performance assessment and risk assessment studies of the NTS.

References

1. Sill, C. Rapid monitoring of soil, water, and air dusts by direct large-area alpha spectrometry. *Health Physics*. 69(1). 21-33. July 1995.
2. Lally, A.E. and Glover, K.E. Source preparation in alpha spectroscopy. *Nuclear Instruments and Methods in Physics Research*. 223. 259-265.