LA-UR-97- - 3 5 0 9

Approved for public release;
distribution is unlimited

Title: Neutron Structural Biology

Author(s): B. P. Schoenborn, LS-DO

Submitted to: DOE Office of Scientific and Technical Information (OSTI)

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

MASTER

Los Alamos

NATIONAL LABORATORY

Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by the University of California for the U.S. Department of Energy under contract W-7405-ENG-36. By acceptance of this article, the publisher recognizes that the U.S. Government retains a nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, make any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Neutron Structural Biology

Benno Schoenborn*
Life Sciences Division

Abstract

This is the final report of a one-year, Laboratory Directed Research and Development (LDRD) project at Los Alamos National Laboratory (LANL). We investigated design concepts of neutron scattering capabilities for structural biology at spallation sources. This included the analysis of design parameters for protein crystallography as well as membrane diffraction instruments. These instruments are designed to be general user facilities and will be used by scientists from industry, universities, and other national laboratories.

Background and Research Objectives

The aim of this project was to investigate design concepts and identify fabrication pitfalls of neutron scattering capabilities for structural biology at spallation sources. This includes the analysis of design parameters for protein and virus crystallography as well as membrane diffraction instruments. These instruments are designed to be general user capabilities that can be used by scientists from industry, universities and other national laboratories. Both hardware and software have to be user friendly but still exhibit the flexibility needed for future as yet undefined experiments.

The use of large detectors will be a major component of any new spectrometer. At present the typical neutron detector is about 17 x 17 cm with a resolution of 1.3 mm at full width at half maximum. Detector concepts will be investigated that allow at least a five-fold increase in active area. For a planned crystallographic application, a detector with a radius of 70 cm and a 20 cm height covering an arc of 120 degrees is envisaged. Such a detector will have a nearly 10-fold increased active area as compared to the presently used detectors.

In addition to such new detectors, new stations will use focusing optics to increase flux. Such optical systems will be based on supermirror technology. Further increases in flux can be achieved by employing moderators matched to the energy resolution needed for the type of experiments. The concept of using such focusing devices has already been successfully tested for vertical focusing but has to be developed for toroidal geometry to approach point focusing.

* Principal Investigator, E-mail: schoenborn@lanl.gov

Importance to LANL's Science and Technology Base and National R&D Needs

The continued development of neutron diffraction and scattering science is a designated LANL thrust area. The development of protein, virus and membrane diffraction capabilities is in direct support of this thrust area. These developments for structural biology will strengthen expertise useful in the development of instruments at the Manuel Lujan, Jr. Neutron Scattering Center (MLNSC) and of similar capabilities at the planned Long Pulse Spallation Source (LPSS) at the Los Alamos Neutron Science Center.

There is a major demand for neutron capabilities for research in structural biology in the US from other national laboratories, universities and industry. There are three neutron scattering instruments for small angle scattering but only one instrument for protein crystallography or membrane diffraction. This station is at Brookhaven National Laboratory (BNL) and is only capable of satisfying a small fraction of the external user demand.

Our completed design of neutron scattering capabilities for structural biology forms the basis for a first neutron instrument that will strengthen the Los Alamos structural biology effort and satisfy the DOE-OBER (Department of Energy-Office of Biological and Environmental Research) goal to bring major diffraction tools to the scientific community. OBER has indicated that it is interested in funding efforts in neutron structural science and supports an external users program.

Scientific Approach and Accomplishments

This one-year task resulted in the completed design of a neutron instrument for application to the structural biology effort at Los Alamos. In addition, this capability will provide a unique resource for a future external users program. The detector has a radius of 70 cm and a 20 cm height covering an arc of 120 degrees. This detector has a nearly 10-fold increased active area as compared to the presently used detectors. In addition to detectors such as the one we have designed, new experimental stations will use focusing optics to increase flux. The optical systems will be based on supermirror technology. Further increases in flux are being studied by employing coupled moderators matched to the energy resolution required.

Publication

1. Schoenborn, B., "A Protein Crystallography Station at LANSCE," in *Neutrons in Biology*, B. P. Schoenborn and R. Knott, Eds. (Plenum Press, New York, 1996) and Los Alamos National Laboratory report LA-UR-96-3508 (November 1996).