OCT 2 1997

SANDIA REPORT

SAND97-2317 • UC-2040

Unlimited Release

Printed September 1997

RECEIVED

OCT 15 1997

OSTI

TDR Calibration for the Alternative Landfill Cover Demonstration (ALCD)

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

Jesus Lopez, Stephen F. Dwyer, James N. Swanson

MASTER

Prepared by
Sandia National Laboratories
Albuquerque, New Mexico 87185 and Livermore, California 94550

Sandia is a multiprogram laboratory operated by Sandia Corporation, a Lockheed Martin Company, for the United States Department of Energy under Contract DE-AC04-94AL85000.

Approved for public release; distribution is unlimited.

Sandia National Laboratories

Issued by Sandia National Laboratories, operated for the United States Department of Energy by Sandia Corporation.

NOTICE: This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government, any agency thereof, or any of their contractors or subcontractors. The views and opinions expressed herein do not necessarily state or reflect those of the United States Government, any agency thereof, or any of their contractors.

Printed in the United States of America. This report has been reproduced directly from the best available copy.

Available to DOE and DOE contractors from
Office of Scientific and Technical Information
P.O. Box 62
Oak Ridge, TN 37831

Prices available from (615) 576-8401, FTS 626-8401

Available to the public from
National Technical Information Service
U.S. Department of Commerce
5285 Port Royal Rd
Springfield, VA 22161

NTIS price codes
Printed copy: A04
Microfiche copy: A01

SAND97-2317
Unlimited Release
Printed September 1997

Distribution
Category UC-2040

TDR Calibration for the Alternative Landfill Cover Demonstration (ALCD)

Jesus Lopez, Stephen F. Dwyer, and James N. Swanson
Environmental Restoration Technologies Department,
Sandia National Laboratories
P.O. Box 5800
Albuquerque, NM 87185-0719

Abstract

The Alternative Landfill Cover Demonstration is a large scale field test that compares the performance of various landfill cover designs in dry environments. An important component of the comparison is the change in the moisture content of the soils throughout the different cover test plots. Time Domain Reflectometry (TDR) is the primary method for the measurement of the volumetric moisture content. Each of the covers is composed of layers of varying types and densities of soils. The probes are therefore calibrated to calculate the volumetric moisture content in each of the different soils in order to gain the optimum performance of the TDR system. The demonstration plots are constructed in two phases; a different probe is used in each phase. The probe that is used in Phase I is calibrated for the following soils: compacted native soil, uncompacted native soil, compacted native soil mixed with six percent sodium bentonite by weight, and sand. The probe that is used in Phase II is calibrated for the following soils: compacted native soil, uncompacted native soil, and sand. In addition, the probes are calibrated for the varying cable lengths of the TDR probes. The resulting empirically derived equations allow for the calculation of in-situ volumetric moisture content of all of the varying soils throughout the cover test plots in the demonstration.

Intentionally Left Blank

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

Contents

Introduction.....	7
Background	9
Soil Measurements With Time Domain Reflectometry	11
Phase I Data Collection	13
6% Bentonite Soil and Compacted Native Soil	13
Uncompacted Native Soil.....	14
Sand.....	14
Phase II Data Collection	15
Sand.....	15
Compacted Native Soil.....	15
Uncompacted Native Soil.....	16
Equation Derivation	17
References	21
Figures	
1 Apparent Length of TDR Probe	11
2 Phase I TDR Probe	12
3 Phase II TDR Probe.....	12
Appendices	
Appendix A: Plots of Third Order Polynomials for Individual Cable Lengths.....	A-1
Appendix B: Volumetric Moisture Content Calculations.....	B-1

Intentionally Left Blank

Introduction

The effectiveness of current EPA landfill cover designs is being questioned, especially in the southwestern states of the United States. The Alternative Landfill Cover Demonstration is a large scale field test to determine the effectiveness of the current accepted designs and compare the results to those of alternative landfill cover designs. The purpose of the demonstration is to determine which cover is the most effective based on cost, ease of construction, and performance. One of the variables being compared in the performance area is the moisture content of the soils in the covers. Time domain reflectometry (TDR) probes measure the moisture content of the soils in each cover. However, an extensive calibration of the probes is required before useful data can be obtained from the probes. Once the calibration is complete, TDR probes provide a means of measuring in-situ volumetric moisture content of soils which allows for a thorough comparison of the various landfill covers.

The following report describes the procedure that is used to calibrate all of the TDR probes used in the Alternative Landfill Cover Demonstration.

Intentionally Left Blank

Background

The Alternative Landfill Cover Demonstration (ALCD), currently underway at Sandia National Laboratories in Albuquerque, New Mexico, is a large scale landfill cover field test. In the demonstration, six landfill cover test plots are constructed side-by-side to compare the plots based on their respective costs, ease of construction, and performance. The construction of the plots is divided into two phases. In Phase I, two covers meeting minimum RCRA Subtitle 'C' and Subtitle 'D' requirements, respectively, are constructed along with Alternative Cover Number one, designed for dry environments. In Phase II, Alternative Covers Two, Three, and Four, which are specifically designed for dry environments, are constructed. The Subtitle 'C' and Subtitle 'D' plots serve as a baseline to compare against the alternative designs in Phase I and Phase II. The results of the field test indicate which of the landfill cover designs is most effective for dry environments.

While the designs of the covers differ, the size and test conditions are the same for all of the test plots. Each of the plots is one hundred meters long and thirteen meters wide; all plots are separated by eleven meters. The plots are situated lengthwise along the east-west axis. Each cover slopes five percent from the middle of the plot fifty meters to the ends of the plot; the layers within each cover design have the same five percent slope. The orientation of the plots allows for passive and active monitoring of the test plots. The western slopes of the plots are monitored under ambient conditions - passive monitoring. A sprinkler system installed in the eastern slopes of the plots allows for stress tests (i.e., simulation of a one hundred year storm) - active monitoring. The similar conditions of the plots allows for an equitable comparison of the cover designs.

All test cover plots are instrumented to quantify water balance variables and ancillary information. For example, various instrumentation installed in and around the covers allows for the continuous collection of the following data: soil moisture status, percolation and interflow, runoff, precipitation, wind speed and direction, relative humidity, and air and soil temperature. The data collection system is automated with manual backup for each individual instrumentation system. Also, periodic measurements will be obtained on vegetation cover, biomass, leaf area index, and species composition. The collection of all such data is tantamount in making a comprehensive comparison of the covers.

The success of the demonstration is heavily dependent on the monitoring systems and care taken prior to and during their installation; as well as follow-on maintenance. One example of care taken prior to the installation is the calibration of the Time Domain Reflectometry (TDR) probes used to determine soil moisture status. TDR is the process of sending pulses through a cable and observing the reflected waveform. If a waveguide, such as a steel rod, is placed on the end of a cable, the reflected waveform is influenced by the dielectric constant of the material surrounding the waveguide. In soils, the dielectric constant of the soil is greatly influenced by the amount of water in the soil; therefore, the reflected waveform from a TDR probe in soil is altered by the moisture content of the soil. The dielectric constant of the soil can be related to the volumetric water constant of the soil in a mathematical expression; thus, a TDR probe can be used to determine in-situ soil moisture content. However, the probes must be carefully calibrated

before they are installed. Different types of soils reflect different waveforms at similar moisture contents. Therefore, a careful calibration of the TDR probes is required before the probes are installed in the test plots in order to ensure the high level of performance required from the monitoring system.

Soil Measurements With Time Domain Reflectometry

Soil moisture measurements are made using TDR, a process of sending pulsed waveforms through a coaxial cable to a fixed length probe and observing the reflected waveform. The reflected waveform is analyzed for two distinct points: the beginning of the probe and the end of the probe. The distance between the two points is identified as the apparent length of the probe (See Figure 1). The apparent length of the probe is proportional to the square root of the apparent dielectric constant of the soil. Because the dielectric constant of water is relatively high in comparison to other medium, a signal will propagate slower in a wetter media (i.e., soil) than in the same medium when dry; therefore, probes in soils with a high moisture content have a longer apparent length than do probes in drier soils. Further, a mathematical relationship between the volumetric moisture content of the soil and the apparent probe length obtained from the waveform can be established. The resulting equation provides a tool for in-situ volumetric soil moisture content measurements using a time domain reflectometry system.

The TDR system that is used in the Alternative Landfill Cover Demonstration (ALCD) consists of Sandia designed and built probes in conjunction with Campbell Scientific's TDR Soil Moisture System for Phase I. Phase II of ALCD uses Campbell Scientific designed and built TDR probes in conjunction with the same CR10 system.

The main components of Campbell Scientific's TDR Soil Moisture System are a Tektronix 1502B cable tester and a CR10 measurement and control system. The use of the 1502B in conjunction with the CR10 allows for the greatest flexibility in the TDR system. The CR10 contains several different algorithms for wave analysis. For example, the CR10 can measure and store entire waveforms, electrical conductivity, or the ratio of apparent length to actual length of the probe. For demonstration purposes, the CR10 is programmed to output the ratio of the apparent length of the rods to the actual length of the rods (L_a/L); volumetric moisture contents are calculated from the ratio in a separate algorithm.

Figure 1. Apparent Length of TDR Probe

The TDR probes for Phase I are a product of extensive research conducted at Sandia National Laboratories. During the design of Phase I, commercially available probes reflect a waveform that can not be used. The waveform does not show a rise when the signal reaches the end of the probe; therefore, the end point of the wave can not be determined and a new probe is required. Components of the probe, such as rod length, rod spacing, number of rods (two or three), coating the rods (with several different types of coatings, coating thickness, and surface preparation), rod diameter, and diode installments are varied until the optimum probe is developed. The resulting probes consist of a potted PVC housing with a length of RG-8 coax cable connected to three equally spaced, six inch long exposed length, epoxy coated prongs (See Figure 2). The center prong is connected to the signal propagating conductor, while the two outside prongs are connected to the coax shield. The epoxy coating inhibits conductivity arising from saline conditions in highly compacted, wet soils. This condition greatly reduces the amplitude of the signal, making meaningful measurements of the time domain difficult to obtain which is a problem in the commercially available probes.

Figure 2. Phase I TDR Probe

When Phase II is in the design phases, Campbell Scientific releases a TDR probe that is capable of reflecting waveforms in the various soils used in the Phase II test plots. In order to conserve the resources used in constructing more probes, the TDR probes for Phase II are designed by and purchased from Campbell Scientific (See Figure 3). The probes consist of a length of RG-8 coaxial cable terminating in an epoxy body. Within the epoxy body, the center conductor of the cable is connected to the center rod of the three parallel 3/16 inch diameter, twelve inch long stainless steel rods. The shield of the coaxial cable is attached to the two outside rods. The rods on the Campbell Scientific probes are not coated due to the increased resolution of the longer rods and the lack of sodium bentonite in the soils producing better defined waveforms.

Figure 3. Phase II TDR Probe

Phase I Data Collection

Various external variables may alter the CR10's analysis of the waveform; therefore, the calibration process requires that the most significant variables be identified and accounted for in the procedure. After several preliminary measurements, two variables are identified to have the greatest impact on the waveform and the CR10's analysis: soil type and cable length. The TDR probes for ALCD, Phase I are placed in four distinct soils: compacted soil mixed with 6% bentonite (by weight), sand, compacted native soil and uncompacted native soil. The CR10 records a different L_a/L ratio for similar moisture contents; therefore, calibration data is collected for each of the different soils in Phase I. Furthermore, the total cable length of each probe affects the waveform and the resulting L_a/L value. After examining the cable lengths that are required for the various soil types in each plot, five cable lengths are identified to adequately represent the entire array of cable lengths; therefore, coaxial cables are prepared to allow for the following lengths: 59', 82', 98', 120', and 138'. In the resulting calibration procedure, data is collected for each of the five cable lengths within each soil type.

To achieve proper soil compaction, a customized mold is designed and manufactured. The mold is similar to a Proctor mold; however, the diameter of the mold is 7.75 inches, and the height is 5.25 inches. Also, a fitting for a TDR probe is built into the mold. The fitting holds the PVC housing of the TDR probes in place while soil is compacted around the probe rods in the mold. The fitting prevents any probe movement and prevents damage that would occur to the probe rod coating if the probe is forced into a compacted brick of soil. A thin PVC pipe is cut to fit within the mold and around the probe. The PVC pipe prevents the metal mold from altering the TDR signal. The custom mold allows for similar compaction and soil conditions to be achieved throughout the entire calibration procedure.

6% Bentonite Soil and Compacted Native Soil

Throughout the calibration procedure, representative field conditions are simulated. Therefore, samples of each of the soils are taken from their respective stockpiles at the construction site and are sieved with a number 4 sieve. The resulting soil is then placed in the mold with four equal lifts. The customized mold is approximately twice the size of the standard Proctor mold; therefore, each lift is delivered 56 blows, more than twice the amount required for a standard Proctor test, with a standard Proctor Compaction hammer. The TDR probe is placed in the mold between the second and third lifts. Once compaction is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. While ten L_a/L values are recorded by the CR10, the soil sample is weighed. Once the ten values are collected, the soil sample is broken apart, and a sample of soil is taken from around the probes and dried to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil is mixed again to achieve a homogeneous sample. The procedure is repeated for five different gravimetric water contents ranging from 1.4% to 18%. After ten L_a/L values are obtained for each of the five different water contents, the cable length is changed and the procedure is repeated for each of the five different cable lengths.

Uncompacted Native Soil

A similar procedure is repeated for the uncompacted native soil. Samples of soil are taken from the stockpile in the field and sieved with a number 4 sieve. The TDR probe is secured into the mold while the soil is placed in the mold and slightly compacted at varying intervals. An inconsistent procedure of compaction simulates field conditions. Once "compaction" is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. Ten L_a/L values are recorded by the CR10, and then the soil sample is weighed. The soil sample is then broken apart, and a sample of soil is taken from around the probes and is dried to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil is mixed again to achieve a homogeneous sample. The procedure is repeated for five different gravimetric water contents ranging from 1.4% to 18%. When L_a/L values are obtained for the five different water contents, the cable length is changed and the procedure is repeated for the five different cable lengths.

Sand

Samples of sand are also taken from the stockpile in the field. The TDR probe is secured into the mold while the sand was placed in the mold and slightly compacted at varying intervals. An inconsistent procedure of compaction is used to simulate field conditions. Once compaction is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. Ten L_a/L values are recorded by the CR10, then the soil sample is weighed. A sample of soil is taken from around the probes and dried to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil is mixed again to achieve a homogeneous sample. The procedure is repeated for five different gravimetric water contents ranging from 1.4% to 18%. When L_a/L values are obtained for the five different water contents, the cable length is changed and the procedure is repeated for the five different cable lengths.

Phase II Data Collection

As in Phase I, soil type and cable length are the most significant variables in affecting the CR10's analysis of the waveform. Therefore, data is collected for the three types of soils in Phase II: sand, compacted native soil, and uncompacted native soil. Also, data is collected for five representative cable lengths, 85, 102, 119, 138, and 156 feet, for each type of soil.

In order to achieve proper compaction of the soil, a mold is designed and built at Sandia National Laboratories to accommodate the larger Phase II probe. The mold is constructed from a one quarter inch steel plate with the following dimensions: 9 inches wide by 24 inches long by 6 inches deep. This size mold allows the entire Phase II probe to lie flat; consequently, a fitting is not required to hold the probe in place. Soil is compacted around the entire probe within the mold.

Sand

Samples of sand are also taken from the stockpile in the field. The TDR probe is secured into the mold while the sand is placed in the mold and slightly compacted at varying intervals. An inconsistent procedure of compaction is used to simulate field conditions. Once compaction is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. Ten L_a/L values are recorded by the CR10, then the soil sample is weighed. A sample of soil is taken from around the probes and is dried in order to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil was mixed again to achieve a homogeneous sample. The procedure is repeated for the following approximate gravimetric water contents: 1.5%, 5%, 7%, 11%, and 15%. For Phase II, the cables used to connect the probes to the 1502B are constructed so that the probe would not have to be removed from the soil when the cable lengths were changed. Therefore, the five moisture contents are the same for all cable lengths within each soil type.

Compacted Native Soil

Throughout the calibration procedure, field conditions are simulated as close as reasonably possible. Therefore, samples of each of the soils are taken from their respective stockpiles at the construction site and are sieved with a number 4 sieve. The resulting soil is then placed in the mold with three equal lifts. As that the customized mold is of an unusual size and shape, tests are conducted to determine the number of blows required from a Standard Proctor hammer in order to achieve 95% compaction; the result is 256 blows per lift. The TDR probe is placed in the mold in the middle of the second lift. Once compaction is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. After ten L_a/L values are recorded by the CR10, the soil sample is weighed. Once the ten values are collected, the soil sample is broken apart, and a sample of soil is taken from around the probes and is dried to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil is mixed again to achieve a homogeneous sample. The procedure is repeated for the following approximate gravimetric water contents: 1.5%, 5%, 7%, 11%, and 15%. For Phase II, the cables used to connect the probes to the 1502B are

constructed so that the probe would not have to be removed from the soil when the cable lengths are changed. Therefore, the five moisture contents are the same for all cable lengths within each soil type.

Uncompacted Native Soil

A similar procedure is repeated for the uncompacted native soil in Phase II as in Phase I. Samples of soil are taken from the stockpile in the field and are sieved with a number 4 sieve. The soil is placed in the mold with three lifts. The TDR probe is placed in the mold in the middle of the second lift. Each lift receives 200 blows with a Standard Proctor Hammer to simulate field conditions. Once "compaction" is complete, the excess soil is shaved from the top of the mold so that the height of soil in the mold is level with the top of the mold. Ten L_a/L values are recorded by the CR10, and then the soil sample is weighed. The soil sample is then broken apart, and a sample of soil is taken from around the probes and dried to determine gravimetric moisture content of the soil per ASTM D4643. Water is then added to the soil, and the soil is mixed again to achieve a homogeneous sample. The procedure is repeated for the following approximate gravimetric water contents: 1.5%, 5%, 7%, 11%, and 15%. For Phase II, the cables used to connect the probes to the 1502B are constructed so that the probe would not have to be removed from the soil when the cable lengths are changed. Therefore, the five moisture contents are the same for all cable lengths within each soil type.

Equation Derivation

The first step in developing the equations is to convert the data into volumetric soil moisture contents (See Appendix B). The gravimetric moisture content, total mass of the sample, and volume of the sample are known variables, the volumetric moisture content is therefore determined from the following equation:

$$\theta_v = \frac{wM_t}{(1+w)V_t}$$

where: θ_v = Volumetric Moisture Content

w = Gravimetric Moisture Content

M_t = Total Mass of Sample

V_t = Total Volume of Sample (Mold)

Once the volumetric moisture content for all of the tests is calculated, the ten L_a/L values recorded by the CR10 are averaged for each moisture level. The five volumetric water contents are plotted against the resulting average L_a/L value for each cable length on an Excel[™] spreadsheet. Using the "Insert Trend Line" feature on Excel, a third order polynomial is found to be the best fit for the data (See Appendix A). After a polynomial for each individual cable length is determined, the data from all of the cable lengths in each soil type is plotted on one graph to verify that separate equations are required for the various cable lengths. Upon examination, the data on the plot containing all of the ratios from the various cable lengths is more scattered than on the plots for individual cable lengths; therefore, the third order polynomial that fit all of the data does not have as high a correlation as the polynomials that fit the data for individual cable lengths. It is therefore apparent that separate polynomials are required for the five different cable lengths in the majority of the soils. The sand, however, only requires two equations: one for the shortest cable and another for the remaining four lengths in Phase I. In Phase II, only one equation is required for sand. The remaining soils require five separate equations. The resulting equations allow for the determination of the volumetric moisture content of the various soils throughout the test plots for the Alternative Landfill Cover Demonstration.

[™] Excel is a registered trademark of Microsoft.

Intentionally Left Blank

Conclusion

TDR probes are a key element in the comparison of the six test plots in the Alternative Landfill Cover Demonstration. However, the probes must be calibrated for their specific type of soil and cable length in order to obtain useful data from the TDR measurements. When calibration is complete, the TDR probes provide a reliable measurement of in-situ volumetric moisture content of the soils that are used throughout the Alternative Landfill Cover Demonstration.

Intentionally Left Blank

References

- ASTM D-4643. Determination of water (moisture) content of soil by microwave oven method.
- Campbell Scientific, Inc. 1992. Addendum to: campbell scientific tdr soil moisture measurement system manual. CR10 Measurement and Control Module Operator's Manual.
- Dasberg, S., and J.W. Hopmans. 1992. Time domain reflectometry calibration for uniformly and nonuniformly wetted sandy and clayey loam soils. Soil Sci. Soc. Am. J. 56:1341-1345.
- Dirksen, C., and S. Dasberg. 1993. Improved calibration of time domain reflectometry soil water content measurements. Soil Sci. Soc. Am. J. 57:660-667.
- Heimovaara, T.J., and W. Bouten. 1990. A computer-controlled 36-channel time-domain reflectometry system for monitoring soil water contents. Water Resour. Res. 27:857-864.

Intentionally Left Blank

Appendix A

Plots of Third Order Polynomials for Individual Cable Lengths

Intentionally Left Blank

6% Bentonite, Phase I, 59' Cable

Figure A-1

6% Bentonite, Phase I, 82' Cable

Figure A-2

Intentionally Left Blank

6% Bentonite, Phase I, 98' Cable

Figure A-3

6% Bentonite, Phase I, 120' Cable

Figure A-4

Intentionally Left Blank

6% Bentonite, Phase I, 138' Cable

Figure A-5

Compacted Native Soil, Phase I, 59' Cable

Figure A-6

Intentionally Left Blank

Compacted Native Soil, Phase I, 82' Cable

Figure A-7

Compacted Native Soil, Phase I, 98' Cable

Figure A-8

Intentionally Left Blank

Compacted Native Soil, Phase I, 120' Cable

Figure A-9

Compacted Native Soil, Phase I, 138' Cable

Figure A-10

Intentionally Left Blank

Uncompacted Native Soil, Phase I, 59' Cable

Figure A-11

Uncompacted Native Soil, Phase I, 82' Cable

Figure A-12

Intentionally Left Blank

Uncompacted Native Soil, Phase I, 98' Cable

Figure A-13

Uncompacted Native Soil, Phase I, 120' Cable

Figure A-14

Intentionally Left Blank

Uncompacted Native Soil, Phase I, 138' Cable

Figure A-15

Sand, Phase I, Short Cables

Figure A-16

Intentionally Left Blank

Sand, Phase I, Long Cables

Figure A-17

Compacted Native Soil, Phase II, 85' Cable

Figure A-18

Intentionally Left Blank

Compacted Native Soil, Phase II, 102' Cable

Figure A-19

Compacted Native Soil, Phase II, 119' Cable

Figure A-20

Intentionally Left Blank

Compacted Native Soil, Phase II, 138' Cable

Figure A-21

Compacted Native Soil, Phase II, 156' Cable

Figure A-22

Intentionally Left Blank

Uncompacted Native Soil, Phase II, 85' Cable

Figure A-23

Uncompacted Native Soil, Phase II, 102' Cable

Figure A-24

Intentionally Left Blank

Uncompacted Native Soil, Phase II, 119' Cable

Figure A-25

Uncompacted Native Soil, Phase II, 138' Cable

Figure A-26

Intentionally Left Blank

Uncompacted Native Soil, Phase II, 156' Cable

Figure A-27

Sand, Phase II, All Cable Lengths

Figure A-28

Intentionally Left Blank

Appendix B

Volumetric Moisture Content Calculations

Intentionally Left Blank

Table B1: Moisture Content for Phase 1 Probe, 138' Cable, 6% Bentonite

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
125.2	122.4	2.29%	15.25	6713.24	150.14	3.70%
209	199	5.03%	15.3	6735.92	322.29	7.94%
125.1	116.6	7.29%	16.81	7420.85	504.21	12.42%
80	71.9	11.27%	18.35	8119.39	822.09	20.26%
198.3	173.5	14.29%	18.8	8323.51	1040.96	25.65%

Table B2: Moisture Content for Phase 1 Probe, 59' Cable, 6% Bentonite

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
157.4	150.8	4.38%	15.08	6636.124467	278.26	6.86%
95.4	89	7.19%	16.38	7225.800599	484.75	11.94%
155.2	140.8	10.23%	16.88	7452.599111	691.48	17.04%
126.5	110.7	14.27%	17.94	7933.411957	990.89	24.42%
209.4	179.3	16.79%	18.59	8228.250023	1182.76	29.14%

Table B3: Moisture Content for Phase 1 Probe, 82' Cable, 6% Bentonite

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
157.4	150.8	4.38%	15.08	6636.12	278.26	6.86%
95.4	89	7.19%	16.38	7225.80	484.75	11.94%
155.2	140.8	10.23%	16.88	7452.60	691.48	17.04%
126.5	110.7	14.27%	17.94	7933.41	990.89	24.42%
209.4	179.3	16.79%	18.59	8228.25	1182.76	29.14%

Table B4: Moisture Content for Phase 1 Probe, 98' Cable, 6% Bentonite

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
197.7	190.4	3.83%	15.52	6835.71	252.41	6.22%
92.7	86.9	6.67%	16.15	7121.47	445.57	10.98%
122.9	112.8	8.95%	16.93	7475.28	614.32	15.14%
132.6	116.9	13.43%	17.75	7847.23	929.12	22.89%
189.8	163.6	16.01%	19.01	8418.76	1162.13	28.64%

Table B5: Moisture Content for Phase 1 Probe, 120' Cable, 6% Bentonite

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
127.4	122.6	3.92%	15.61	6876.53	259.08	6.38%
108.6	102	6.47%	15.65	6894.67	419.01	10.32%
107.4	98	9.59%	16.4	7234.87	633.22	15.60%
125	110.8	12.82%	17.29	7638.57	867.74	21.38%
191.4	164.9	16.07%	18.1	8005.99	1108.46	27.31%

Table B6: Moisture Content for Phase 1 Probe, 59' Cable, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
145.7	143.7	1.39%	14.5	6577.16	90.28	2.22%
70.7	67.4	4.90%	15.6	6871.99	320.76	7.90%
110.1	102.3	7.62%	17.22	7606.82	538.90	13.28%
117.4	106.3	10.44%	18.56	8214.64	776.68	19.14%
231.6	200.2	15.68%	18.3	8096.71	1097.74	27.05%

Table B7: Moisture Content for Phase 1 Probe, 82' Cable, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
145.7	143.7	1.39%	14.5	6577.16	90.28	2.22%
70.7	67.4	4.90%	15.6	6871.99	320.76	7.90%
110.1	102.3	7.62%	17.22	7606.82	538.90	13.28%
117.4	106.3	10.44%	18.56	8214.64	776.68	19.14%
231.6	200.2	15.68%	18.3	8096.71	1097.74	27.05%

Table B8: Moisture Content for Phase 1 Probe, 98' Cable, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
139.5	135	3.33%	14.31	6490.97	209.39	5.16%
95	89.2	6.50%	15.85	6985.39	426.48	10.51%
126.9	115.2	10.16%	17.74	7842.69	723.09	17.82%
212.8	191.8	10.95%	18.43	8155.67	804.84	19.83%
170	146.6	15.96%	18.27	8083.10	1112.61	27.42%

Table B9: Moisture Content for Phase 1 Probe, 120' Cable, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
160.7	157.7	1.90%	14.6	6622.52	123.63	3.05%
86	82	4.88%	14.93	6568.08	305.49	7.53%
130.1	120.9	7.61%	16.68	7361.88	520.59	12.83%
152.2	136.5	11.50%	18.26	8078.56	833.33	20.53%
217.5	188.6	15.32%	18.16	8033.20	1067.40	26.30%

Table B10: Moisture Content for Phase 1 Probe, 138' Cable, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
189.2	185.8	1.83%	14.66	6649.73	119.50	2.94%
84.5	80.4	5.10%	14.99	6595.30	320.01	7.89%
143.7	133	8.05%	16.66	7352.81	547.50	13.49%
149.8	134.3	11.54%	18.34	8114.85	839.65	20.69%
213.8	185.2	15.44%	18.49	8182.89	1094.62	26.97%

Table B11: Moisture Content for Phase 1 Probe, 59' Cable, Sand

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
183.7	181.9	0.99%	15.28	6726.84	65.91	1.62%
204.5	197.7	3.44%	15	6599.84	219.46	5.41%
194.8	183.3	6.27%	14.73	6477.37	382.39	9.42%
271.7	244.3	11.22%	15.99	7048.90	710.86	17.52%
297.8	249.9	19.17%	17.09	7547.85	1214.04	29.91%

Table B12: Moisture Content for Phase 1 Probe, 82' Cable, Sand

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
183.7	181.9	0.99%	15.28	6726.84	65.91	1.62%
204.5	197.7	3.44%	15	6599.84	219.46	5.41%
194.8	183.3	6.27%	14.73	6477.37	382.39	9.42%
271.7	244.3	11.22%	15.99	7048.90	710.86	17.52%
297.8	249.9	19.17%	17.09	7547.85	1214.04	29.91%

Table B13: Moisture Content for Phase 1 Probe, 138' Cable, Sand

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
161.8	159.9	1.19%	15.47	6813.03	80.00	1.97%
141.1	136.7	3.22%	14.5	6373.04	198.73	4.90%
223.1	209.9	6.29%	15.33	6749.52	399.34	9.84%
192.4	175.8	9.44%	15.35	6758.60	583.12	14.37%
284.4	253.7	12.10%	15.89	7003.54	756.01	18.63%
279.7	234	19.53%	16.8	7416.31	1211.75	29.86%

Table B14: Moisture Content for Phase 1 Probe, 59' Cable, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
136.8	134.8	1.48%	13.68	6205.21	90.72	2.24%
175.6	167.8	4.65%	12.29	5574.71	247.62	6.10%
96	90.4	6.19%	12.09	5483.99	319.90	7.88%
118.2	108.7	8.74%	14.83	6726.84	540.65	13.32%
178.9	160.1	11.74%	15.68	7112.40	747.42	18.42%
246.9	213.7	15.54%	16.83	7634.04	1026.53	25.29%

Table B15: Moisture Content for Phase 1 Probe, 82' Cable, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
136.8	134.8	1.48%	13.68	6205.21	90.72	2.24%
175.6	167.8	4.65%	12.29	5574.71	247.62	6.10%
96	90.4	6.19%	12.09	5483.99	319.90	7.88%
118.2	108.7	8.74%	14.83	6726.84	540.65	13.32%
178.9	160.1	11.74%	15.68	7112.40	747.42	18.42%
246.9	213.7	15.54%	16.83	7634.04	1026.53	25.29%

Table B16: Moisture Content for Phase 1 Probe, 98' Cable, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
160.2	156.4	2.43%	13.67	6200.67	147.08	3.62%
84.4	80	5.50%	13.3	6032.84	314.51	7.75%
158.3	145.7	8.65%	12.56	5697.18	453.47	11.17%
114.3	102.5	11.51%	15.85	7189.51	742.22	18.29%
159.7	139	14.89%	15.63	7089.72	918.96	22.64%

Table B17: Moisture Content for Phase 1 Probe, 120' Cable, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
127.3	125.1	1.76%	13.46	6105.42	105.51	2.60%
110.2	104.6	5.35%	13.11	5946.66	302.19	7.45%
126	115.1	9.47%	13.29	6028.30	521.50	12.85%
209.8	186.4	12.55%	14.94	6776.74	755.84	18.62%
208.3	180.1	15.66%	16.51	7488.89	1013.86	24.98%

Table B18: Moisture Content for Phase 1 Probe, 138' Cable, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
83.9	80.3	4.48%	11.5	5216.37	223.82	5.52%
199.6	186.8	6.85%	12.48	5660.89	363.02	8.95%
127	115.5	9.96%	13.72	6223.35	563.53	13.89%
145.8	128.5	13.46%	16.14	7321.06	868.68	21.40%
208.7	175.3	19.05%	16.86	7647.65	1223.92	30.16%

Table B19: Moisture Content for Phase 2 Probe, All Cables, Compacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
170.1	168	1.25%	78.63	35667	440.33	2.15%
379.5	362.6	4.66%	78.57	35640	1587.13	7.75%
224.8	210.8	6.64%	85.41	38743	2412.82	11.77%
596	536.3	11.13%	93.48	42892	4296.40	20.97%
248.2	215.9	14.96%	95.6	42823	5572.86	27.20%

Table B20: Moisture Content for Phase 2 Probe, All Cables, Sand

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
146.3	144.3	1.39%	78.76	35800	489.41	2.39%
229.4	219.3	4.61%	76.15	34615	1524.03	7.44%
252.8	236.3	6.98%	80.00	36365	2373.51	11.58%
214.7	194.2	10.56%	84.77	38531	3679.02	17.95%
344.3	297.9	15.58%	89.89	40858	5506.28	26.87%

Table B21: Moisture Content for Phase 2 Probe, All Cables, Uncompacted Native Soil

M of Wet Soil	M of Solids	Moisture Content (w)	Total Mass (lbs)	Total Mass (g)	V of Water	Volumetric w
136.4	134.7	1.26%	74.74	33971	423.39	2.07%
148.7	142.3	4.50%	71.40	32455	1396.85	6.82%
218.3	204.4	6.80%	76.51	34775	2214.26	10.81%
189	171.5	10.20%	85.79	38997	3610.83	17.62%
256.9	222	15.72%	92.04	41837	5683.58	27.74%

DISTRIBUTION:

1 Willis Stewart
AGEC Inc.
2019 Butler Loop
Richland, WA 99352	1 Jim Studer
Intera, Inc.
1650 University Blvd. N.E.
Albuquerque, NM 87102
1 Ralph Gruebel
AGRA Earth & Environmental, Inc.
4700 Lincoln Rd., N.E.
Albuquerque, NM 87109	1 Cindy Ardito
Intera, Inc.
1650 University Blvd. N.E.
Albuquerque, NM 87102
1 Brent Lewis
Bureau of Land Management
Denver Service Center
Bldg. 50
Denver, CO 80225-0047	1 Joe Shinn
Lawrence Livermore National Laboratory
MS L453
P.O. Box 808
Livermore, CA 94551-9900
1 Stephanie Odell
Bureau of Land Management
Farmington District Office
1235 LaPlata Highway
Farmington, NM 87401	1 Earl Whitney
Los Alamos National Laboratory
EES-5, MS F665
Los Alamos, NM 87545
1 Larry Stebbins
FERMCO
MS: 81-2
P.O. Box 538704
Cincinnati, OH 45235	1 Kay Birdsell
Los Alamos National Laboratory
EES-5, MS F665
Los Alamos, NM 87545
1 Kevin Kostelnick
Idaho National Engineering &
Environmental Lab
Lockheed Martin Idaho Technologies
P.O. Box 1625
Idaho Falls, ID 83415-3710	1 Rosenberg Nina
Los Alamos National Laboratory
EES-DO, MS D446
Los Alamos, NM 87545
1 Wayne Downs
Idaho National Engineering &
Environmental Lab
P.O. Box 1625
Idaho Falls, ID 83415-2110	1 Steven Limback
Los Alamos National Laboratory
ESA-EPE, MS J576
Los Alamos, NM 87545
	1 Bruce Erdal
Los Alamos National Laboratory
EM-TD, MS J591
Los Alamos, NM 87545

- 1 Eric Rogoff
MDM/Lamb, Inc.
6121 Indian School Rd., N.E.
Suite 105
Albuquerque, NM 87110
- 1 Andrea Hart
MSE Technology Applications, Inc.
P.O. Box 4078
Butte, MT 59702
- 1 Creighton Barry
MSE Technology Applications, Inc.
P.O. Box 4078
Butte, MT 59702
- 1 Dave Emilia
MSE Technology Applications, Inc.
P.O. Box 4078
Butte, MT 59702
- 1 Robert Andrews, Ph.D.
National Research Council
National Academy of Science
Board of Radioactive Waste Management
2001 Wisconsin Ave., N.W.
Washington, DC 20007
- 1 New Mexico Junior College
Pannell Library
Lovington Highway
Hobbs, NM 88240
- 1 New Mexico State Library
325 Don Gaspar
Santa Fe, NM 87503
- 1 New Mexico Tech
Martin Speere Memorial Library
Campus Street
Socorro, NM 87810
- 1 J. Espinosa
NM Environment Department
1190 St. Francis Drive
Santa Fe, NM 87503-0968
- 1 Elizabeth Fiedler
Oak Ridge National Laboratory
P.O. Box 2008
Oak Ridge, TN 37831-6317
- 1 Tom Early
Oak Ridge National Laboratory
P.O. Box 2008
Oak Ridge, TN 37831-6317
- 1 Cindy Kendrick
Oak Ridge National Laboratory
P.O. Box 2001, EW-923
Oak Ridge, TN 37831
- 1 Kirk Cantrell
Pacific Northwest National Laboratories
MS K6-81
P.O. Box 999
Richland, WA 99352
- 1 Phillip Gauglitz
Pacific Northwest National Laboratories
Battelle Blvd.
P.O. Box 999
Richland, WA 99352
- 1 Steve Slate
Pacific Northwest National Laboratories
Battelle Blvd.
P.O. Box 999
Richland, WA 99352
- 1 Sandy Dalvit-Dunn
Science & Engineering Assoc., Inc.
3205 Richards Lane
Suite A
Santa Fe, NM 87505

1 Bill Lowry
Science & Engineering Assoc., Inc.
3205 Richards Lane
Suite A
Santa Fe, NM 87505

1 Andrew Murray
SubTerra Remedial Engineers, Inc.
1820 Bering Drive
Suite 23
San Jose, CA 95112-4212

1 Cheryl Allen
Tech Reps, Inc.
5000 Marble N.E.
Albuquerque, NM 87110

1 D. Dresp
Thomas Branigan Library
106 W. Hadley St.
Las Cruces, NM 88003

1 Marvin Gross
U.S. Department of Energy
Fernald Field Office
P.O. Box 538704
Cincinnati, OH 45253-8704

1 Johnny Moore
U.S. Department of Energy
Oak Ridge Operations Office
EW-923
P.O. Box 2001
Oak Ridge, TN 37831-8620

1 Elisabeth Reber-Cox
U.S. Department of Energy
Oakland Operations Office
Room 700 N
1301 Clay Street
Oakland, CA 94612-5208

1 Helen Stolz
U.S. Department of Energy
Nevada
P.O. Box 98518
Las Vegas, NV 89193-8518

1 Gillian Eaton
U.S. Department of Energy
Rocky Flats
Tech Site, Bldg. T124A
P.O. Box 928
Golden, CO 80402-0928

1 Shannon Saget
U.S. Department of Energy
Richland Operations Office
P.O. Box 550, K8-50
Richland, WA 99352

1 Skip Chamberlain
U.S. Department of Energy, EM-53
Cloverleaf Building
19901 Germantown Rd.
Germantown, MD 20874-1290

1 Mel Shupe
U.S. Department of Energy
Federal Energy Technology Center
Industrial Park
P.O. Box 3462
Butte, MT 59702

1 Bob Bedick
U.S. Department of Energy
Federal Energy Technology Center
3610 Collins Ferry Rd.
Morgantown, WV 26507-0880

1 Dale Pflug
U.S. Department of Energy
Chicago Operations Office
EAD/900
9800 South Cass Ave.
Argonne, IL 60439-4832

1 Doug Maynor
U.S. Department of Energy
Ohio Operations Office
P.O. Box 3020
Miamisburg, OH 45343-3020

1 Thomas Williams
U.S. Department of Energy
Idaho Operations Office
MS1219
785 DOE Pl.
Idaho Falls, ID 83402

1 Richard Baker
U.S. Department of Energy
Chicago Operations Office
9800 South Cass Ave.
Argonne, IL 60439

1 Paul Beam
U.S. Department of Energy, EM-40
Cloverleaf Building
19901 Germantown Rd.
Germantown, MD 20874-1290

1 Kim Abott
U.S. Department of Energy
Oakland Operations Office
Room 700 N
1301 Clay Street
Oakland, CA 94612-5208

1 Dennis Green
U.S. Department of Energy
Idaho Operations Office
850 Energy Drive
Idaho Falls, ID 83401-1563

1 Dennis Olona
U.S. Department of Energy
Albuquerque Operations Office
P.O. Box 5400
Albuquerque, NM 87185

1 Paul Zielinski
U.S. Department of Energy, EM-443
Cloverleaf Bldg.
19901 Germantown Rd.
Germantown, MD 20874-1290

1 James Wright
U.S. Department of Energy
Savannah River Operations Office
Bldg. 703-46A
P.O. Box A
Aiken, SC 29803

1 Bill Wilborn
U.S. Department of Energy
Nevada Operations Office
2753 S Highland Rd.
Las Vegas, NV 89109

1 Phillip Washer
U.S. Department of Energy
Savannah River Operations Office
Bldg. 773-A
P.O. Box 616
Aiken, SC 29803

1 Jef Walker
U.S. Department of Energy, EM-53
Cloverleaf Bldg.
19901 Germantown Rd.
Germantown, MD 20874-1290

1 George Schneider
U.S. Department of Energy
Idaho Operations Office
785 DOE Place
Idaho Falls, ID 83402

1 Elizabeth Phillips
U.S. Department of Energy
Oak Ridge Operations Office
P.O. Box 2001
Oak Ridge, TN 37830

<p>1 Rod Warner
U.S. Department of Energy
Ohio Operations Office
P.O. Box 538705
Cincinnati, OH 45030</p>	<p>1 James Paulson
U.S. Department of Energy
Chicago Operations Office
9800 South Cass Ave.
Argonne, IL 60439</p>
<p>1 Scott McMullin
U.S. Department of Energy
Savannah River Operations Office
Bldg. 703-46A
P.O. Box A
Aiken, SC 29802</p>	<p>1 Gary Huffman
U.S. Department of Energy
Rocky Flats Office
Highway 93rd & Cactus St.
Golden, CO 80402</p>
<p>1 Julianne Levings
U.S. Department of Energy
Albuquerque Operations Office
P.O. Box 5400
Albuquerque, NM 87185</p>	<p>1 Sharon Johnson
U.S. Department of Energy
Savannah River Operations Office
703 A, Rm. B202
Aiken, SC 29802</p>
<p>1 Jeff Lenhart
U.S. Department of Energy
Albuquerque Operations Office
P.O. Box 5400
Albuquerque, NM 87185</p>	<p>1 John Geiger
U.S. Department of Energy
Savannah River Operations Office
P.O. Box A
Aiken, SC 29802</p>
<p>1 Chuck Morgan
U.S. Department of Energy
Nevada Operations Office
2753 S. Highland Rd.
Las Vegas, NV 89109</p>	<p>1 Arturo Palomares
U.S. Environmental Protection Agency
Region 8
999 18th Street
Suite 500
Denver, CO 80202</p>
<p>1 Thomas Hicks
U.S. Department of Energy
Savannah River Operations Office
Bldg. 703-46A
P.O. Box A
Aiken, SC 29802</p>	<p>1 John Stormont
University of New Mexico
Dept. of Civil Engineering
Albuquerque, NM 87131-1351</p>
<p>1 Maria Vargas
U.S. Department of Energy
Richland Operations Office
P.O. Box 550
Richland, WA 99352</p>	<p>1 University of New Mexico
Zimmerman Library
Government Publications Department
Albuquerque, NM 87131</p>
	<p>1 Bruce Thompson
University of New Mexico
Dept. of Civil Engineering
Albuquerque, NM 87131-1351</p>

1	David Daniel University of Texas at Austin Dept. of Civil Engineering Cockrell Hall, Room 9.102 Austin, TX 78712	1	MS0701	Richard Lynch, 6100
		1	0719	George Allen, 6621
		25	0719	Thomas Burford, 6621
		1	0719	Stephen Dwyer, 6621
		1	0719	Jesus Lopez, 6621
		1	0719	J. Christopher Sears, 6621
1	Thomas Gaughan Westinghouse Savannah River Company Savannah River Technology Center Environmental Sciences Section P.O. Box 616 / 730-2B Aiken, SC 29802	1	0724	Joan Woodard, 6000
		1	0968	James Swanson, 5716
		1	1147	Warren Cox, 6681
		1	1147	Fran Nimick, 6682
		1	1148	Dick Fate, 6685
		1	1148	Jerry Peace, 6685
		2	0619	Review & Approval Desk for DOE/OSTI, 12690
1	Michael Serrato Westinghouse Savannah River Company Savannah River Technology Center Environmental Sciences Section P.O. Box 616 / 773-42A Aiken, SC 29802	5	0899	Technical Library, 4916
		1	9018	Central Technical Files, 8940-2
1	Tim Jarosch Westinghouse Savannah River Company Bldg. 773-42A Road SR-1 Aiken, SC 29808			
1	Pete Zionkowski Westinghouse Savannah River Company Savannah River Technology Center Environmental Sciences Section P.O. Box 616 / 730-2B Aiken, SC 29802			
1	Brian Looney Westinghouse Savannah River Company Bldg. 773-42A Road SR-1 Aiken, SC 29808			
1	Ahmet Seur Westinghouse Savannah River Company Savannah River Technology Center Environmental Sciences Section P.O. Box 616 / 730-2B Aiken, SC 29802			