CONF-970706--

ERNEST ORLANDO LAWRENCE BERKELEY NATIONAL LABORATORY

New Schemes in the Adjustment of Bendable, Elliptical Mirrors Using a Long Trace Profiler

RECEIVED
OCT 16 1997
OSTI

Seungyu Rah, Scott Locklin, Steve Irick,
and M.R. Howells

**Accelerator and Fusion
Research Division**

MASTER

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

August 1997

Presented at the
*International Conference
on Optical Science,
Engineering and
Instrumentation 1997,*
San Diego, CA,
July 28 - August 1, 1997,
and to be published in
SPIE Proceedings

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor The Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or The Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof, or The Regents of the University of California.

Ernest Orlando Lawrence Berkeley National Laboratory
is an equal opportunity employer.

New Schemes in the Adjustment of Bendable, Elliptical Mirrors Using a Long Trace Profiler

Seungyu Rah¹, Scott Locklin², Steve Irick², M.R. Howells²

¹Beamline Division
Pohang Light Source/POSTECH
Pohang, 790-784, Korea

²Advanced Light Source
Ernest Orlando Lawrence Berkeley National Laboratory
University of California, Berkeley, California 94720

DISCLAIMER

**Portions of this document may be illegible
in electronic image products. Images are
produced from the best available original
document.**

New schemes in the adjustment of bendable, elliptical mirrors using a Long Trace Profiler

Seungyu Rah
Beamline Division
Pohang Light Source/POSTECH
Pohang, 790-784, Korea

Scott Locklin, Steve Irick, and Malcolm Howells
University of California,
Lawrence Berkeley National Laboratory
Advanced Light Source
Berkeley, CA 94720, USA

ABSTRACT

The Long Trace Profiler (LTP), an instrument for measuring the slope profile of long X-ray mirrors, has been used for adjusting bendable mirrors. Often an elliptical profile is desired for the mirror surface, since many synchrotron applications involve imaging a point source to a point image. Several techniques have been used in the past for adjusting the profile measured in height or slope of a bendable mirror. Underwood et al. have used collimated X-rays for achieving a desired surface shape for bent glass optics¹. Non linear curve fitting using the simplex algorithm was later used to determine the best fit ellipse to the surface under test². A more recent method uses a combination of least squares polynomial fitting to the measured slope function in order to enable rapid adjustment to the desired shape.

The mirror has mechanical adjustments corresponding to the first and second order terms of the desired slope polynomial, which correspond to defocus and coma, respectively. The higher order terms are realized by shaping the width of the mirror to produce the optimal elliptical surface when bent. The difference between desired and measured surface slope profiles allows us to make methodical adjustments to the bendable mirror based on changes in the signs and magnitudes of the polynomial coefficients. This technique gives rapid convergence to the desired shape of the measured surface, even when we have no information about the bender, other than the desired shape of the optical surface. Nonlinear curve fitting can be used at the end of the process for fine adjustments, and to determine the over all best fit parameters of the surface. This technique could be generalized to other shapes such as toroids.

1. INTRODUCTION

As manufacturing techniques developed, fabrication of aspherical surfaces became a well established technique in the optics industry and aspherical optical components became widely used for the visible wavelength range to get high quality imaging performance for a variety of fields of application from automatic cameras to large aperture telescopes. However, in the x-ray region, because of reflection characteristics and the short wavelength nature of the light, highly polished grazing incidence reflective optics are required. Reflective optics are sensitive to surface error and more so to short wavelengths. Therefore, high quality aspherical surfaces are required. It is difficult to get high quality aspherical mirror surfaces for the X-ray region by grinding and polishing to the desired figure. However, it is easier to get a high quality plane surface and bend it into the desired aspherical profile. A well known example is the bendable monolithic mirror which gives a circular cylinder³. A microspot X-ray photon spectroscopy (μ XPS) beamline recently under construction at the Advanced Light Source (ALS) uses two elliptically bent cylindrical mirrors in a Kirkpatrick-Baez configuration to get a beam with a spot size of about $1\text{ }\mu\text{m} \times 1\text{ }\mu\text{m}$. The similarity between the

bending equation for a uniform beam and polynomial expansion of a mirror surface shape gives us insight into how aspheric surfaces can be realized by bending. The mirror bender is designed to mechanically adjust the first and second order terms of mirror slope function using unequal couples⁴. The width of the mirrors have a specially designed edge shape to achieve the desired high order terms of the mirror profile when bent. For the optimal bending of the mirrors, the Long Trace Profiler(LTP)^{5, 6} has been used during the bending process. For quick and accurate bending of the mirror, an exact equation representing the aspheric surface is derived and used.

2. BENDING THEORY

When an elastic beam experiences unequal couples C_1 and C_2 (these are torques shown in Figure 1), the bending equation of the beam is expressed as follows⁴:

$$\frac{d^2 z}{dy^2} = \frac{1}{E \cdot I(y)} \left[\frac{C_1 + C_2}{2} + \frac{C_1 - C_2}{L} y \right] \quad (1)$$

where E is the Young's modulus, $I(y)$ is the section moment, and L is the length of the beam. The surface height profile of any cylinder can be represented as a polynomial equation:

$$\frac{d^2 z}{dy^2} = 2a_2 + 6a_3 y + 12a_4 y^2 + 20a_5 y^3 + \dots \quad (2)$$

Equations (1) and (2) are similar, especially when the mirror has a large radius of curvature. In this case the terms higher than the cubic term in Equation (2) are negligibly small. Also, by adjusting the unequal couples C_1 and C_2 , the zeroth and first order terms of Equation (1) can be controlled to match equation (2). For optimal matching between Equations (1) and (2), the section moment function $I(y)$ must change with respect to position y when the mirror is bent. This can be achieved either by varying the thickness of the mirror or by varying the width of the mirror. The mirrors that described here are examples of changing mirror width.

Figure 1. Schematic drawing of an elliptically bent cylindrical mirror with unequal couples.

3. REPRESENTATION OF A CONIC CYLINDER AS A POWER POLYNOMIAL AND THE EXACT EQUATION

Many synchrotron beamline designs involve imaging a point source to a point image, and thus require an ellipsoidal mirror surface. Special cases where the source point is at infinity or the image point is virtual will require a parabolic or hyperbolic surface. This section will describe the exact equation and its expansion coefficients for the conic surfaces which include ellipses, parabolas, hyperbolas, and circles.

Given the source distance r , image distance r' , and the incident angle of the principal ray on the mirror measured from the surface normal θ , the ellipsoidal mirror shown in Figure 2 will have foci at $(0, -r \sin \theta, r \cos \theta)$ and $(0, r' \sin \theta, r' \cos \theta)$. The equation of the surface is represented as follows by the definition of an ellipsoid⁷.

$$z = \frac{x^2 \sec^2 \theta + y^2}{R_0 \{ (1 - Ay) + \sqrt{(1 - Ay)^2 - B(x^2 \sec^2 \theta + y^2)} \}} \quad (3)$$

In this equation the origin of the coordinates is at the intersection of the principal ray and the mirror. The parameters A , B , and R_0 are defined as follows:

$$\frac{\cos^2 \theta}{r} + \frac{\cos^2 \theta}{r'} = \frac{2 \cos \theta}{R_0} \quad (4)$$

$$A = \frac{\sin \theta}{2} \left(\frac{1}{r'} - \frac{1}{r} \right) \quad (5)$$

$$B = A^2 + \frac{1}{rr'} \quad (6)$$

Figure 2. Coordinates for the representation of elliptical mirror.

In Equations (4), (5), and (6), r , r' and θ have positive values. If either r or r' has a negative value, then Equation (3) represents a hyperboloid. When either r or r' is set to infinity, it represents a paraboloid. When $r = r'$ and $\theta = 0$, then it is the equation of a sphere. Setting $x = 0$ in Equation (3) will yield a conic cylinder, which could be any of an elliptical, hyperbolic, parabolic, or circular cylinder.

A power expansion of Equation (3) for the cylindrical mirror surface gives the coefficients a_i of Equation(2)

$$a_2 = \frac{1}{2R_0}, \quad (7) \quad a_3 = a_2 A, \quad (8)$$

$$a_4 = a_2 \frac{4A^2 + B}{4}, \quad (9) \quad a_5 = a_2 A \frac{4A^2 + 3B}{4}, \quad (10)$$

$$a_6 = a_2 \frac{8A^4 + 12A^2B + B^2}{8}, \quad (11) \quad a_7 = a_2 A \frac{8A^4 + 20A^2B + 5B^2}{8}, \quad (12)$$

$$a_8 = a_2 \frac{64A^6 + 240A^4B + 120A^2B^2 + 5B^3}{64} \quad (13)$$

$$a_9 = a_2 A \frac{64A^6 + 336A^4B + 280A^2B^2 + 35B^3}{64} \quad (14)$$

$$a_{10} = a_2 \frac{128A^8 + 896A^6B + 1120A^4B^2 + 280A^2B^3 + 7B^4}{128} \quad (15)$$

.....

Since the LTP measures the slope of a mirror, we represent the mirror profile with slope equations. Taking the derivative of Equation (3) gives the following slope function of a conic cylinder:

$$\frac{\partial z}{\partial y} = \frac{1}{BR_0} \left\{ \frac{A - A^2y + By}{\sqrt{(1 - Ay)^2 - By^2}} - A \right\}. \quad (16)$$

4. DESCRIPTION OF THE μ XPS MIRRORS

By using two elliptically bent cylindrical mirrors in a Kirkpatrick-Baez configuration, the beamline designers intend to get a spot in the image plane having a size of $1 \mu\text{m} \times 1 \mu\text{m}$. The source is an elliptically shaped pinhole with dimensions $40 \mu\text{m}$ (horizontal) $\times$ $20 \mu\text{m}$ (vertical), and the parameters of the two mirrors are listed in Table 1.

The mirrors are made of 17-4PH stainless steel⁸ and have constant thickness. They have a specially designed edge shape in order for the mirror to have an elliptical surface when it is bent. To get the $1 \mu\text{m}$ spot size (2σ width), it is necessary to adjust the mirrors into the desired surface shape with a maximum of 0.97 and $2.1 \mu\text{rad}$ RMS of surface slope error, respectively. To get the proper performance of the mirrors, ± 0.01 mm of width accuracy and ± 0.1 mm thickness accuracy are required. Also, the back and front surface should be parallel to better than ± 0.002 mm. Given the long source distance, short image distance, and long mirror length, the mirrors' surfaces correspond to that part of the ellipse where there is much curvature change over the length of the mirror. A few lower order terms in the polynomial expansion of the surface shape are not enough to represent the desired shape within the acceptable slope error. Therefore, either more polynomial terms or an exact equation must be used. Before bending, the mirrors were installed on the mirror bender and the surfaces were measured by both the LTP and a ZYGO interferometer with a150

Mirror	Parameters	Aperture size
M1 (Vertical Focusing Mirror)	$r = 3880mm$ $r' = 220mm$ $\theta = 88.4$	$80mm(L) \times 2mm(W)$
M2 (Horizontal Focusing Mirror)	$r = 4000mm$ $r' = 100mm$ $\theta = 88.4$	$50mm(L) \times 2mm(W)$

Table 1. Parameters of the mirrors bent.

mm diameter aperture. The mirrors have errors for polishing flat mirror. Mirror M1 (clear aperture of 80 mm (L) $\times$ 5 mm (W)) has 3.5 μ rad RMS slope error and mirror M2 (clear aperture of 50 mm (L) $\times$ 3 mm (W)) has 3.7 μ rad RMS slope error when they are flat.

5. BENDING OF MIRRORS

As is mentioned previously, there are two adjustments for controlling C_1 and C_2 . When the section moment $I(y)$ is a slowly varying function of position y , then the coefficients a_2 and a_3 have a close relationship with C_1 and C_2 as follows:

$$2a_2 = \frac{1}{2EI}(C_1 + C_2) \quad (17)$$

$$6a_3 = \frac{1}{EIL}(C_1 + C_2) \quad (18)$$

The Equations (17) and (18), gives us enough information for a rough adjustment of C_1 and C_2 , in the bending procedure. However, Figures 3 and 4 shows that the two terms are not enough to explain the mirror shape, for the final stage of bending. Equation (19) below defines a function to check the bending error as the difference D_{sl} between measured slope $(\frac{dz}{dy})_m$ and the desired slope $\frac{dz}{dy}$ that was represented in Equation (16):

$$D_{sl} = (\frac{dz}{dy})_m - \frac{dz}{dy} \quad (19)$$

Equation (19) gives us much information about the mirror being bent. The Equation (16) is exact equation, representing a conic cylinder which could be a parabola, hyperbola, ellipse or circle. And, the Equation (19) will not give us term truncation errors for the calculation of the desired mirror surface. The shape, or the polynomial expansion coefficient of the Equation (19), gives the signed magnitude for the adjustment of couples, similar to equations (17) and (18). The RMS value of the function D_{sl} evaluated over the length of the mirror has a close relationship with image quality of the mirror. Use of equation (19) is therefore needed to get the mirror surfaces with least RMS slope error, by adjusting the couples C_1 and C_2 , further more. For these reason, even if there is no information about the amount of required couples for the desired shape of the mirror, it is still possible to get the right shape of mirror if the mirror shape parameters are known. By the iterative application of adjustments, measurement and analysis of surface slope, the bending process shows good convergence and the desired curve can be achieved within an acceptable amount of error boundary as long as the mechanical adjustment systems are working properly.

Figure 3. Surface height truncation error of mirror M1 after taking up to 3rd, 4th, and 5th order terms of polynomial expansion.

After bending the mirrors, mirror M1 has 3.1 μrad RMS of slope error for a 76 mm long clear aperture and mirror M2 has 2.5 μrad RMS of slope error for a 40 mm long clear aperture. The slope errors before and after bending are similar. The slope errors for each case are shown in Figures 5 and 6. These figures clearly show the expected "S" shape, which predicts that spherical aberration will be the dominant aberration in the image. The mirrors are installed in the beamline and show a beam spot of approximately 2 μm (horizontal) x 3 μm (vertical).

Figure 4. Surface height truncation error of mirror M2 after taking up to 3rd, 4th, and 5th order terms of polynomial expansion.

Figure 5. Slope error of mirror M1, after bending. The RMS error is $3.1 \mu\text{rad}$ for a 76 mm aperture.

Figure 6. Slope error of mirror M2, after bending. The RMS error is $2.5 \mu\text{rad}$ for a 40 mm aperture.

6. CONCLUSION

The long trace profiler, an instrument measuring the slope profile of mirrors, has been used for adjusting bendable mirrors into elliptical cylinders. In order to quickly adjust the mirrors, exact equations representing a conic surface shape and its derivatives are derived. Comparing the slope from the equation with the slope from the measurement gives us directional and quantitative information for bending the mirror. By iterative application of bending and measuring the mirror, the desired mirror shape is achieved. It is possible to bend mirrors into the desired shape even if the mechanical characteristics of the mirror bender are not known.

ACKNOWLEDGMENTS

The authors thank Greg Morrison and Keith Frank for their help in setting up the bender mechanisms and for constantly adding improvements.

This work was supported by the Director, Office of Energy Research, Office of Basic Energy Sciences, Materials sciences Division of the U. S. Department of Energy, Under contract No. DE-AC03-76SF00098.

REFERENCES

1. J. H. Underwood and D. Turner, "Bent glass optics", *Proc. SPIE*, Vol. 106, pp. 125-135, 1977.
2. S. C. Irick, "Pre-adjustment of small, elliptical bender mirrors for an X-ray beamline", *Proc. SPIE*, Vol. 2515, pp. 37-41, July, 1995.
3. M. R. Howells, "Design Strategies for Monolithic Adjustable-Radius Metal Mirrors", *Opt. Eng.*, Vol. 34, No. 2, pp 410-417, 1995.
4. H. A. Padmore, M. R. Howells, S. Irick, T. Renner, R. Sandler, and Y. M. Koo, "Some new schemes for producing high-accuracy elliptical X-ray mirrors by elastic bending", *Proc. SPIE*, Denver, CO Aug, 1996.
5. P. Z. Takacs and S. Qian, *United States Patent* 4884697, 1989.
6. S. C. Irick, W. R. McKinney, D. L. J. Lunt, and P. Z. Takacs, "Using a straightness reference in obtaining more accurate surface profiles from a long trace profiler," *Review of Scientific Instruments*, Vol. 63, No.1 (Part IIB), pp. 1436-1438, January, 1992.
7. Seungyu Rah and Malcolm R. Howells, "Representation of conic surfaces and conic cylinders with vertex coordinates system for design and assessment of grazing incident optics", *LBL internal report*, 1997.
8. M. R. Howells and R. A. Paquin, "Optical substrate materials for synchrotron radiation beamlines", *LBNL internal report*, LSBL #388, 1997.