**Cover Sheet for a Hanford
Historical Document
Released for Public Availability**

Released 1995

**Prepared for the U.S. Department of Energy
under Contract DE-AC06-76RLO 1830**

**Pacific Northwest Laboratory
Operated for the U.S. Department of Energy
by Battelle Memorial Institute**

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, make any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

DECLASSIFIED

By Authority of

DOC

May 1973

By

J. E. Savely 11-16-94

Verified By

J. E. Savely 11-17-94

SPECIAL RE-REVIEW
FINAL DETERMINATION
DECLASSIFICATION CONFIRMED

BY JP Derouin DATE 5-5-81
BY JW Jordan DATE 5-5-81

RECORD CENTER FILE

2. [redacted]
3. P. W. [redacted]
4. H. W. Murray
5. R. E. Roberts
6. I. F. Butler
7. D. F. Shepard
8. A. H. Bushey
9. G. B. Barton
10. 300 File
11. 700 File
12. Pink Copy
13. Yellow Copy

July 14, 1950

This document consists of
3 Pages No. [redacted]

O. P. Amacker
2704 L Building - 200 E Area

RE: ANALYSIS OF A STANDARD PU SOLUTION
BY THE CONTROL LABORATORIES IN 231 and 234-5 BUILDINGS

A chemically pure solution of Pu was prepared for use in a re-investigation of the present 49 titration method. The purification of this solution consisted of an oxalate precipitation from a diluted Pu solution followed by dissolution in nitric acid. Two peroxide precipitations were carried out with the final precipitate being dissolved in a small volume of concentrated HNO_3 .

Spectrographic analysis of this sample indicated the following impurities which are expressed as p.p.m.: Ca 40, Cr 40, La 40, Mg 4, Na 10, Ni 20.

Three 500 μ l. portions of the sample were dried and ignited to PuO_2 . From the weight of the residue and the impurity analysis, the concentration, g/L, of the solution was found to be 222.2, 221.5, and 222.5; average 222.0.

RECEIVED

JUN 1 1956

300 AREA
CLASSIFIED FILES

MASTER

REPRODUCED FROM BEST
AVAILABLE COPY

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DECLASSIFIED

This standardization was carried out April 13. On May 11 six samples of this solution were submitted to the control laboratory in 231. On May 23 six samples of this solution were submitted to the control laboratory in 234-5. The results reported were:

<u>231 Laboratory</u>	<u>234-5 Laboratory</u>
219.3	228.2
223.6	233.3
221.5	220.2
220.9	236.1
222.3	225.2 re-run 220.6
228.2 re-run 229.3	227.9
average 222.6	average 227.3

The analyses made by the control laboratories were not corrected for iron since this constituent is known to be low in these samples.

Since a considerable lapse of time occurred between the gravimetric standardization and submission of the samples to the laboratories for

analysis, a correction for increase in concentration due to decomposition of the water by alpha particles and evaporation was calculated. Four 10 ml. flasks of the solution had been weighed at the time of filling and again after 54 days to give a rate of loss of water from the solution. The calculated concentrations at time of submission of the samples were 222.4 g/l for 231 laboratory samples and 222.6 g/l for the 234-5 laboratory samples. A correction for iron equivalent to approximately 0.4 g/l should be applied to the laboratory results.

The results reported by the 231 laboratory are very consistent excepting for one determination and the average value agrees with the gravimetric determination. The average of results from the 234-5 laboratory are two

DECLASSIFIED 18319

-3-

percent higher than the average of the gravimetric results and the method does not appear to be under as good control as it is in the 231 laboratory.

Further studies are being made on the chemical method for plutonium to solve the difficulty.

G. B. Barton

G. B. Barton
Analytical Section
Technical Services Division

GBB:kb