RECEIVED
JUN 19 2000
OSTI

**X-ray Scattering Facility
for
Dynamics of Surfaces and Interfaces**

Final Report

For Period Aug. 15, 1997---Aug. 15, 1999

E. W. Plummer

The University of Tennessee
Department of Physics and Astronomy
401 A. H. Nielsen Physics Building
Knoxville, TN 37996-1200

Aug. 31, 1999

Prepared for

We have no objection from a patent
standpoint to the publication or
dissemination of this material.

MP Driscoll
Office of Intellectual
Property Counsel
DOE Field Office, Chicago

6/16/2000
Date

THE U. S. DEPARTMENT OF ENERGY
AWARD NO. DOE DE-FG02-96ER45586
Funding \$744 K

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, make any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

Review of goals

This project has been funded by the Department of Energy to allow the principal investigators, working with the Complex Materials Consortium Collaborative Access Team (CMC-CAT), to build a facility for the study of surfaces and interfaces using x-rays at the Advanced Photon Source (APS) at Argonne National Laboratory. Research at this facility will be directed toward determination of structure, dynamics, phase transitions and growth in complex systems including surfaces, thin films and interfaces, magnetic surfaces and alloys. The CMC-CAT, representing eight institutions from industry, national laboratories, and universities, has obtained permission from APS to construct beamlines at the insertion device and bending magnet (this project) of Sector 9.

The partners in this project are 1) Exxon, 2) The University of Tennessee, 3) Oak Ridge National Laboratory, and 4) the University of Pennsylvania. The funding for the project is listed below.

Exxon	\$500K
Penn	\$100K
UTK	\$100K
ORNL	\$250K
DOE	\$744 K

CMC-CAT Bending Magnet Beamline

The CMC-CAT bending magnet beamline in Sector 9 at the Advanced Photon Source was designed with two major classes of experiments in mind: UHV surface diffraction and EXAFS. The layout of this line is shown in the two accompanying pictures. Besides other features the first class requires an optical arrangement that faithfully preserves the energy-diffraction angle relation in a collimated, non-dispersive setting, while the second requires focussed beam with high energy resolution. A combination of a fast, high-resolution monochromator succeeded by a toroidal mirror was chosen as a compromise to best satisfy both classes.

The beamline configuration features a differential pump as the first optical element inside the First Optics Enclosure (FOE) 9-BM A. It accepts an incident fan of synchrotron radiation of 2 mrad horizontal width and, since it makes a solid window as transition between the storage ring and the beamline vacuum obsolete, enables a usable energy range that reaches down to the phosphor absorption edge at about 2.1 keV.

The next optical elements are a white beam slit to limit the incident beam divergence and a water-cooled, double crystal monochromator. The monochromator is specifically designed to support an extended energy range from 2.1 keV to about 30 keV and enable fast energy scans for Quick-EXAFS measurements. The energy resolution is determined

by the choice of diffracting crystals and a fast, convenient way to change these crystals for specific experiments will be implemented.

The incident white beam is stopped by a water-cooled copper absorber, which passes the monochromatic beam with a vertical offset of 25 mm.

Beam-focussing and rejection of higher harmonics is accomplished through a bendable toroidal mirror, which can be set to focus in either the EXAFS experimental station 9-BM B or the surface diffraction station 9-BM C. The mirror is rhodium coated and will allow reflection angles of up to 3.5 mrad. For non-dispersive diffraction set-ups a second flat mirror will be provided in the B-station.

A monochromatic slit further defines the divergence of the monochromatic beam and the photon shutter near the down-stream end of the FOE isolates the experimental stations from the FOE.

The monochromatic beam is transported through a line of integrally shielded beampipes into the experimental stations 9-BM B (EXAFS) and 9-BM C (Surface Diffraction). Space is available in three pump station enclosures to place beam diagnostic or beam conditioning devices as needed.

UHV surface diffraction experiments will be conducted inside the hutch using a surface surface chamber and diffractometer donated by Exxon.

A detailed account of the purchases made with the DOE funds is attached.

9-BM First Optics Enclosure (FOE)

CMC-CAT Bending Magnet Beamline