**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 2 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 348 THRU 692)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INORGANIC ANALYSIS, DIFFERENTIAL SCANNING CALORIMETRY ANALYSIS, DSC, THERMOGRAVIMETRIC ANALYSIS, TGA, SPECIFIC GRAVITY ANALYSIS, SPG		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 8
of 17Sec 9
of 17Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

HNF-1663 REV. 0

INORGANIC ANALYSIS

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

LABCORE Data Entry Template for Worklist# 26781

Analyst: KA Instrument: DSC0 3 Book # 12/14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.56</u>	<u>N/A</u>	Joule:/g
98000436	BY-105 (2)	~~3 SAMPLE~~	~~S98T002392~~	0	DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joule:/g
98000436	BY-105 (2)	3 DUP	S98T002392	0	DSC-03	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joule:/g
98000437	BY-105 (2)	4 SAMPLE	S98T002806	0	DSC-03	LIQUID	<u>N/A</u>	<u>13.44</u>		Joule:/g
98000437	BY-105 (2)	5 DUP	S98T002806	0	DSC-03	LIQUID	<u>13.44</u>	<u>11.58</u>	<u>N/A</u>	Joule:/g

Final page for worklist # 26781

Neil King 10/22/98
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Validated 11/11/98 BBachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND103001 Fri Oct 30 08:38:58 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 351 TO 353.

Curve 1: DSC
 File Info: SAM103007 Fri Oct 30 12:41:22 1998
 Sample Weight: 12.340 mg
 998Y002898

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 14:08:11 1998

Curve 1: DSC
 File Info: SAK103C08 Fri Oct 30 13:36:51 1998
 Sample Weight: 14.640 mg
 SSK1003805 DHP

LABCORE Data Entry Template for Worklist# 26782

Analyst: SLH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>27.64</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	2 SAMPLE	S98T002807	0	DSC-03	LIQUID	<u>N/A</u>	<u>11.94</u>		Joules/g
98000437	BY-105 (2)	3 DUP	S98T002807	0	DSC-03	LIQUID	<u>11.94</u>	<u>12.52</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	4 SAMPLE	S98T002808	0	DSC-03	LIQUID	<u>N/A</u>	<u>9.658</u>		Joules/g
98000437	BY-105 (2)	5 DUP	S98T002808	0	DSC-03	LIQUID	<u>9.658</u>	<u>10.88</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26782

Janet L. Hood
Analyst Signature
10-30-98
Date

Kim Thomas
Analyst Signature
11-10-98
Date

Validated 11/11/98 Q. Nachelov

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File info: IND102901 Thu Oct 29 10:17:54 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 355 TO 359.

N2, EXOTHERM DOWN
 140.0 °C
 180.0 °C
 0.0 min RATE: 10.0 °C/min
 TIME: Thu Oct 29 22:05:32 1998

JI SOLBRACK
 PERKIN-ELMER
 7 Series Thermal Analysis System

Jeffrey Beatty

Curve 1: DSC

File Info: SAM102912 Thu Oct 29 20:35:50 1998

Sample Weight: 15.340 mg

S98T002807

N2 10C/min

TEMP: 35.0 °C

TEMP: 400.0 °C

TIME: 0.0 min

RATE: 10.0 °C/min

Temperature (°C)

SL HOOD BOATRIGT
PERKIN-ELMER7 Series Thermal Analysis System
Mon Nov 2 13:45:26 1998

Curve 1: DSC

File Info: SAM102913 Thu Oct 29 22:32:35 1998

Sample Weight: 16.290 mg

S98T002807 DUP

Curve 1: DSC
 File info: SAM102914 Thu Oct 29 23:27:15 1998
 Sample Weight: 16.820 mg
 S98T002808

N2 10C/min
 TEMP: 35.8 °C
 TIME: 00:00
 RATE: 0.0 min RATE: 10.0 C/min
 SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Nov 2 13:52:20 1998

Curve 1: DSC
 File Info: SAM102915 Fri Oct 30 00:20:51 1998
 Sample Weight: 13.950 mg
 S98T002808 DUP

worklistrpt Version 2.1 05/15/95
10/23/98 11:08

Page: 1

LABCORE Data Entry Template for Worklist# 26783 Analyst: JIS Instrument: DSC0 3 Book # 12N14BMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DI.	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>27.13</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	2 SAMPLE	S98T002809	0	DSC-03	LIQUID	<u>N/A</u>	<u>15.90</u>		Joules/g
98000437	BY-105 (2)	3 DUP	S98T002809	0	DSC-03	LIQUID	<u>15.90</u>	<u>26.86</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	4 SAMPLE	S98T002810	0	DSC-03	LIQUID	<u>N/A</u>	<u>22.05</u>		Joules/g
98000437	BY-105 (2)	5 DUP	S98T002810	0	DSC-03	LIQUID	<u>22.05</u>	<u>25.26</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26783

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File Info: IND102901 Thu Oct 29 10:17:54 1998

Sample Weight: 19.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 361 TO 365.

JT SOLBRACK Jeff Solbrack
PERMAN-ELMER
7 Series Thermal Analysis System
Thu Oct 29 10:19:09 1998

Curve 1: DSC
 File info: SAM102908 Thu Oct 29 12:13:22 1998
 Sample Weight: 13.270 mg
 S98T002809

N2 10C/min
 TEMPERATURE 85.0 °C
 TIME: 0.0 min RATE: 10.0 C/min

Temperature (°C)

JI SOLBRACK
 PEAKAN-CLIMED
 7 Series Thermal Analysis System
 Mon Nov 2 15:24:37 1998

Curve 1: DSC
 File info: SAM102909 Thu Oct 29 13:45:23 1998
 Sample Weight: 13.990 mg
 S98T002809DUP

Curve 1: DSC

File info: SAM102910 Thu Oct 29 14:49:06 1998

Sample Weight: 14.090 mg

S98T002810

Curve 1: DSC

File Info: SAM102911 Thu Oct 29 15: 46: 10 1998

Sample Weight: 13.860 mg

S98T002810DUP

N2 100./min

TEMP: 35.0 °C

TIME: 666.8 s

TIME: 0.0 min RATE: 10.0 °C/min

Temperature (°C)

JI SOLBRACK
PERKIN-ELMER7 Series Thermal Analysis System
Mon Nov 2 15: 32: 52 1998

LABCORE Data Entry Template for Worklist# 26784

Analyst: SLH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.41</u>	<u>28.12</u>	<u>N/A</u>	Joulee/g
98000437	BY-105 (2)	2 SAMPLE	S98T002811	0	DSC-03	LIQUID	<u>N/A</u>	<u>4.41</u>		Joules/g
98000437	BY-105 (2)	3 DUP	S98T002811	0	DSC-03	LIQUID	<u>4.41</u>	<u>4.79</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	4 SAMPLE	S98T002812	0	DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g
98000437	BY-105 (2)	5 DUP	S98T002812	0	DSC-03	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26784

Andrew H. Hood
Analyst Signature Date

MLB
Analyst Signature Date

Validated 11-3-98
Rejected 11/2/98 Michael

598T002811

Data Entry Comments:

598T002812 integrated improperly. Reintegration not possible due to unsaved data. Rerun requested. MB 11/9/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File info: IND110302 Tue Nov 3 06:49:54 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 367 TO 369.

Sandra Hood Bostright

Curve 1: DSC
 File info: qsav2 Tue Nov 3 07:43:30 1998
 Sample Weight: 16.390 mg
 698700230-S440 11-3-98
 598700812

N2 10C/min
 TEMPERATURE 300.0 °C
 TIME: 0.0 min RATE: 10.0 °/min
 SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 08:40:31 1998

Curve 1: DSC
 File info: SAM110302 Tue Nov 3 09:33:47 1998
 Sample Weight: 15.380 mg
 S98T002812 DUP

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 09:45:47 1998

N2 10C/min
 10.0 C/min
 0.0 min RATE15
 TIME2:
 000.0 g

LABCORE Data Entry Template for Worklist# 26785

Analyst: PM Instrument: DSC0 3 Book # 12N143Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				DSC-03	LIQUID	<u>28.45</u>	<u>28.14</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	2	SAMPLE	S9BT002813	0		DSC-03	LIQUID	<u>N/A</u>	<u>2.635</u>		Joules/g
98000437	BY-105 (2)	3	DUP	S9BT002813	0		DSC-03	LIQUID	<u>2.635</u>	<u>1.437</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	4	SAMPLE	S9BT002938	0		DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	5	DUP	S9BT002938	0		DSC-03	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26785

Analyst Signature PM Date 11/2/98Analyst Signature Kim Thomas Date 11-5-98Validated 11/10/98 JS

Data Entry Comments:

High RPD S9BT002813 due to sample inhomogeneities and small size of peak.Rerun only at customer request. JS 11/10/98Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

Curve 1: DSC
 File Info: IND110201 Mon Nov 2 19:08:38 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 371 TO 375.

N2, EXOTHERM DOWN

140.0 °C
 180.0 °C
 0.0 min RATE: 10.0 °C/min

Temperature (°C)

RD MEYERS
 PERKIN-ELMER
 / Series Thermal Analysis System
 Mon Nov 2 19:26:59 1998

371

Curve 1: DSC
 File info: SAM110203 Mon Nov 2 20:22:42 1998
 Sample Weight: 13.160 mg
 S98T002813

N2 10C/min
 250.0 °C
 12.00 g

0.0 min PWT: 10.0 u/min

Temperature (°C)

RD MEYERS
 PERKIN-ELMER

7 Series Thermal Analysis System
 Tue Nov 3 14:48:27 1998

Curve 1: DSC
 File Info: SAM110202 Mon Nov 2 21:17:48 1998
 Sample Weight: 13.260 mg
 S98T002813DUP

N2 10C/min

TEMP: 28.9 °C TIME: 0.0 min RATE: 10.0 °C/min

RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 14:51:32 1998

Curve 1: DSC
 File Info: SAM110204 Mon Nov 2 22:55:36 1998
 Sample Weight: 14.040 mg
 S98T002938

Curve 1: DSC
 File Info: SAM10206 Tue Nov 3 00:05:13 1998
 Sample Weight: 15.050 mg
 S98T002938DUP

RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 14:59:33 1998

N2 10C/min
 800.0 g
 0.0 min RATE
 10.0 C/min

LABCORE Data Entry Template for Worklist# 26786

Analyst: DK Instrument: DSC0 3 Book # 12714-8

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.22</u>	N/A	Joules/g
98000509	BY-105 (2)	2 SAMPLE	S98T002957	0	DSC-03	LIQUID	N/A	<u>4.99</u>		Joules/g
98000509	BY-105 (2)	3 DUP	S98T002957	0	DSC-03	LIQUID	<u>4.99</u>	<u>15.89</u>	N/A	Joules/g
98000509	BY-105 (2)	4 TRIPLE	S98T002957	0	DSC-03	LIQUID	<u>4.99</u>	<u>14.62</u>		Joules/g
98000509	BY-105 (2)	5 SAMPLE	S98T002958	0	DSC-03	LIQUID	N/A	<u>11.76</u>		Joules/g
98000509	BY-105 (2)	6 DUP	S98T002958	0	DSC-03	LIQUID	<u>11.76</u>	<u>9.18</u>	N/A	Joules/g
98000509	BY-105 (2)	7 SAMPLE	S98T002959	0	DSC-03	LIQUID	N/A	<u>13.19</u>		Joules/g
98000509	BY-105 (2)	8 DUP	S98T002959	0	DSC-03	LIQUID	<u>13.19</u>	<u>12.90</u>		Joules/g
							<u>13.19</u>	<u>12.90</u>	<u>10/26/98</u>	Joules/g

Final page for worklist # 26786

Analyst Signature Mike Kim Date 10/24/98

Analyst Signature Kim Thomas Date 10-29-98

Validated 10/29/98 Drachler

Data Entry Comments:

High RPD S98T002957 due to sample inhomogeneities. Triplicate run in good agreement with duplicate. No rerun. Q33 10/26/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND102301 Fri Oct 23 20: 45: 42 1998

Sample Weight: 11.670 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 377 TO 384.

Curve 1: DSC

File Info: SAM102303 Sat Oct 24 00:08:18 1998

Sample Weight: 11.700 mg

S98T002957

HNF-1663 REV. 0

Curve 1: DSC
File info: SAM102401 Sat Oct 24 01:23:43 1998
Sample Weight: 16.900 mg
S98T002957 DUP

Curve 1: DSC
 File info: SAM102402 Sat Oct 24 02:24:43 1998
 Sample Weight: 15.490 mg
 S98T002957 TRIP

Curve 1: DSC
 File info: SAM102301 Fri Oct 23 22:03:56 1998
 Sample Weight: 15.340 mg
 S98T002958

Curve 1: DSC

File Info: SAM102302 Fri Oct 23 22:56:08 1998

Sample Weight: 13.040 mg

S98T002958 DUP

Curve 1: DSC

File info: SAM102403 Sat Oct 24 04:17:58 1998

Sample Weight: 16.630 mg

S98T002959

HNF-1663 REV. 0

Curve 1: DSC
 File Info: SAM102404 Sat Oct 24 05:12:55 1998
 Sample Weight: 15.640 mg
 S98T002959 DUP

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 14:03:19 1998

LABCORE Data Entry Template for Worklist# 26787

Analyst: RDM Instrument: DSC0 3 Book # 12N14B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.69</u>	<u>N/A</u>	Joules/g
98000509	BY-105 (2)	2 SAMPLE	S98T002955	0	DSC-03	LIQUID	<u>N/A</u>	<u>27.66</u>		Joules/g
98000509	BY-105 (2)	3 DUP	S98T002955	0	DSC-03	LIQUID	<u>27.66</u>	<u>33.76</u>	<u>N/A</u>	Joules/g
98000509	BY-105 (2)	4 SAMPLE	S98T002956	0	DSC-03	LIQUID	<u>N/A</u>	<u>29.50</u>		Joules/g
98000509	BY-105 (2)	5 DUP	S98T002956	0	DSC-03	LIQUID	<u>29.50</u>	<u>29.16</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26787

RDM 10/31/98
Analyst Signature Date

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/11/98 Bachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File Info: IND103001 Fri Oct 30 08:39:58 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 386 TO 390.

N2, EXOTHERM DOWN

TEMP: 140.0 °C
TEMP: 180.0 °C

0.0 MIN RATE: 10.0 °/MIN

Temperature (°C)

RM KING

DEPT. CLERK

7 Series Thermal Analysis System
Tue Nov 3 15:08:58 1998

Curve 1: DSC

File Info: SAM103101 Sat Oct 31 00:46:57 1998

Sample Weight: 13.510 mg

S98T002955

N2 10C/min

TEMP: 500.0 °C

TIME:

0.0 min RATE:

40.0 °C/min

Temperature (°C)

AD MEYERS

DEKATON-ELMER

7 Series Thermal Analysis System
Tue Nov 3 16:12:58 1998

Curve 1: DSC
 File Info: SAM103102 Sat Oct 31 01:50:25 1998
 Sample Weight: 15.060 mg
 S98T002955DUP

N2 10C/min

TEMP: 555.0 °C

TIME: 6

0.0 min RATE: 10.0 °C/min

Temperature (°C)

RD MEYERS
 DUNN CLARK

7 Series Thermal Analysis System
 Tue Nov 3 15:18:14 1998

Curve 1: DSC
 File info: SAM103103 Sat Oct 31 02:40:38 1998
 Sample Weight: 13.710 mg
 S98T002956

N2 10C/min
 TEST: 565-18 8
 TIME: 0.0 min HATES: 20.0 C/min
 RD MEYERS
 PERKIN ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 15:24:08 1998

Curve 1: DSC
 File info: SAM103104 Sat Oct 31 03:40:37 1998
 Sample Weight: 15.470 mg
 S98T002956DUP

RD MEYERS
 PERMAN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 15:26:50 1998

N2 100C/min
 THERM: 55.0 °C
 TEMPE: 550.0 °C
 TIME: 0.0 min
 RATE: 10.0 °C/min

LABCORE Data Entry Template for Worklist# 26788

Analyst: SLH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.04</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T002589	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000436	BY-105 (2)	3 DUP	S98T002589	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	4 SAMPLE	S98T002590	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000436	BY-105 (2)	5 DUP	S98T002590	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26788

Sandra Howard Boatright
Analyst Signature Date

10-28-98

Validated 11/3/98 SLH

Kim Thomas 10-29-98
Analyst Signature Date

Data Entry Comments:

Ran trip on S98T002589 & S98T002590
11/3/98

Second STD run due to delay in completion of worklist with a value of 28.19 J/g
and a recovery of 99.09%. S98T002589

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File Info: IND102702 Tue Oct 27 16:53:44 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 392 TO 397.

N2, EXOTHERM DOWN

TEMP: 180.0 °C TIME: 0.0 min RATE: 10.0 °C/min

Temperature (°C)

SL HOOD BOATRIGHT
PERKIN-ELMER

7 Series Thermal Analysis System
Wed Oct 28 04:36:09 1998

Jeffrey Boatright

Curve 1: DSC

File Info: SAH102803 Wed Oct 28 03:58:15 1998

Sample Weight: 25.900 mg

S98T002589

N2 10C/min
TIME: 260.0 g
TIME: 260.0 g

SL HOOD BOATRIGT
PEAKIN-ELMER
7 Series Thermal Analysis System
Wed Oct 28 04:03:59 1998

Curve 1: DSC

File info: SAM102804 Wed Oct 28 04:55:45 1998

Sample Weight: 23.390 mg

S98T002589 DUP

Curve 1: DSC

File info: IND102801 Wed Oct 28 18:37:14 1998

Sample Weight: 11.670 mg

STD 12N14-B

N2, EXOTHERM DOWN

TIME: 148.88 TIME: 148.88

0.0 min RATE: 10.0 C/min

SL HOOD BOATRIGT

PERKIN-ELMER

7 Series Thermal Analysis System

Wed Oct 28 18:48:06 1998

Indur-Hood Boatright

Curve 1: DSC
 File Info: SAM102807 Wed Oct 28 20:46:59 1998
 Sample Weight: 21.640 mg
 9987002590-DUP-Samp 8/2/99

SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 20:46:45 1998

Curve 1: DSC
 File info: SAM1092808 Wed Oct 28 21:41:27 1998
 Sample Weight: 23.710 mg
 S98T002590 TTP-DuP 03/10/25/98

SL HODD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 22:31:21 1998

LABCORE Data Entry Template for Worklist# 26789

Analyst: JH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.12x</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T002591	0	DSC-03	SOLID	<u>N/A</u>	<u>6.27</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	3 DUP	S98T002591	0	DSC-03	SOLID	<u>6.27</u>	<u>5.63</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	4 SAMPLE	S98T002592	0	DSC-03	SOLID	<u>N/A</u>	<u>1.59</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	5 DUP	S98T002592	0	DSC-03	SOLID	<u>1.59</u>	<u>1.70</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26789

Analyst Signature Sandra Hood Boatright
Date 11-3-98

Analyst Signature McBry
Date 11/9/98

Validated 11/9/98 McBry

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND110302 Tue Nov 3 06:48:54 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 399 TO 403.

X1 154.600 °C
X2 160.466 °C
Peak 157.251 °C
Area 366.156 mJ
ΔH 28.123 J/g
Height 37.242 mW
Onset 155.157 °C

Heat Flow (mW)

N2, EXOTHERM DOWN
TEMP 148.8 °C
TIME 1.0

0.0 min RATE: 10.0 C/min

Temperature (°C)

SL HOOD BOATRIGT
PERKIN-ELMER

7 Series Thermal Analysis System
Tue Nov 3 06:50:36 1998

Handwritten signature

Curve 1: DSC

File Info: SAM110305 Tue Nov 3 13:42:01 1998

Sample Weight: 17.230 mg

S98T002591

SL HOOD BOATRIGHT
PERKIN-ELMER
7 Series Thermal Analysis System
Mon Nov 9 05:46:30 1998

Curve 1: DSC

File Info: SAM10306 Tue Nov 3 15: 16: 39 1998

Sample Weight: 24.590 mg

S98T002591 DUP

Curve 1: DSC

File Info: SAM110307 Tue Nov 3 16:10:50 1998

Sample Weight: 23.250 mg

s98t002952

Curve 1: DSC

File info: SAM110308 Tue Nov 3 17:47:07 1998

Sample Weight: 25.420 mg

S98T002592 DUP

N2 10C/min
TEMP 300.0 °C
TIME: 00.0 g

0.0 min RATE: 10.0 C/min

SL HOOD BOATRIGT
PERKIN-ELMER

7 Series Thermal Analysis System

Mon Nov 9 09:54:00 1998

LABCORE Data Entry Template for Worklist# 26790

Analyst: SH Instrument: DSC0 3 Book # 12114-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>27.40</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T002593	0	DSC-03	SOLID	<u>N/A</u>	<u>246.6</u>		Joules/g
98000436	BY-105 (2)	3 DUP	S98T002593	0	DSC-03	SOLID	<u>246.6</u>	<u>213.6</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	4 SAMPLE	S98T002594	0	DSC-03	SOLID	<u>N/A</u>	<u>170.0</u>		Joules/g
98000436	BY-105 (2)	5 DUP	S98T002594	0	DSC-03	SOLID	<u>170.0</u>	<u>227.1</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26790

Sandra Howard Boatright
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Validated 11/12/98 Shachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File Info: IND102901 Thu Oct 29 10:17:54 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 403 TO 409.

N2, EXOTHERM DOWN

140.0 g TIMES: 0.0 min RATE: 10.0 °C/min

Temperature (°C)

JT SOLBRACK
PERKIN-ELMER

7 Series Thermal Analysis System
Fri Oct 30 01:44:21 1998

Settled Bentley

Curve 1: DSC
 File info: SAM103001 Fri Oct 30 01:57:43 1998
 Sample Weight: 23.760 mg
 S98T002593

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Nov 2 14:00:55 1998

Curve 1: DSC

File Info: SAM103002 Fri Oct 30 02:51:18 1998

Sample Weight: 16.870 mg

S98T002593 DUP

N2 10C/min

TEMP: 98.0 °C

TIME: 000.0

RATES: 0.0 min RATE: 10.0 °C/min

Temperature (°C)

SL HOOD BOATRRIGHT

PERKIN-ELMER

7 Series Thermal Analysis System

Mon Nov 2 14:03:28 1998

Curve 1: DSC
 File Info: SAM103003 Fri Oct 30 04:34:33 1998
 Sample Weight: 18.440 mg
 S98T002594

Curve 1: DSC
 File info: SAM103004 Fri Oct 30 05:29:11 1998
 Sample Weight: 24.090 mg
 S98T002594 DUP

N2 10C/min

TEMP1: 35.0 °C
 TEMP2: 500.0 °C

TIME: 0.0 min RATE: 10.0 °C/min

SL HOOD BOATRIGHT
 PERMIN-ELMER
 7 Series Thermal Analysis System
 Mon Nov 2 14:51:27 1998

LABCORE Data Entry Template for Worklist# 26791

Analyst: Rpm Instrument: DSC0 3 Book # 12M43Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.63*</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T002711	0	DSC-03	SOLID	<u>N/A</u>	<u>20.76</u>		Joules/g
98000436	BY-105 (2)	3 DUP	S98T002711	0	DSC-03	SOLID	<u>20.76</u>	<u>17.72</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	4 SAMPLE	S98T002737	0	DSC-03	SOLID	<u>N/A</u>	<u>18.19</u>		Joules/g
98000490	BY-105 (2)	5 DUP	S98T002737	0	DSC-03	SOLID	<u>18.19</u>	<u>20.86</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26791

Analyst Signature Rpm Date 10/31/98Analyst Signature Kim Thomas Date 11-10-98Validated 11/11/98 J. Pacheco

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND103001 Fri Oct 30 08:39:58 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 411 TO 415.

N2, EXOTHERM DOWN

TEMP: 140.0 °C
TEMP2: 180.0 °C

0.0 MIN RATE: 10.0 °/min

Temperature (°C)

RM KING

DEPT: CLM

7 Series Thermal Analysis System

Tue Nov 3 15:08:58 1998

Signature

Curve 1: DSC

File info: SAM103017 Fri Oct 30 22:21:28 1998

Sample Weight: 17.790 mg

S98T002711

N2 10C/min

TEMP: 800.0 °C

TAP: 0.5

GAS FLOW RATE: 30.0 mL/min

RD MEYERS

PERKIN-ELMER

7 Series Thermal Analysis System
Tue Nov 3 16:04:57 1998

Curve 1: DSC
 File info: SAM03018 Fri Oct 30 23:13:17 1998
 Sample Weight: 17.860 mg
 S98T002711DUP

Curve 1: DSC
 File info: SAM103015 Fri Oct 30 19:42:21 1998
 Sample Weight: 29.550 mg
 S98T002737

RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 19:55:05 1998

N2 10C/min
 38.0 g
 500.0 g

TIME: 1

Curve 1: DSC

File Info: SAM103016 Fri Oct 30 20:47:08 1998

Sample Weight: 23.380 mg

S98T002737DUP

LABCORE Data Entry Template for Worklist# 26792

Analyst: KRM Instrument: DSC0 3 Book # 124MB

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD.			DSC-03	SOLID	<u>28.45</u>	<u>28.12</u>	N/A	Joules/g
98000437	BY-105 (2)	2 SAMPLE	S98T002840	0	DSC-03	SOLID	N/A	<u>29.84</u>		Joules/g
98000437	BY-105 (2)	3 DUP	S98T002840	0	DSC-03	SOLID	<u>29.84</u>	<u>23.86</u> <u>25.89</u>	<u>14.27</u> N/A	Joules/g
98000437	BY-105 (2)	4 SAMPLE	S98T002841	0	DSC-03	SOLID	N/A	<u>28.56</u>		Joules/g
98000437	BY-105 (2)	5 DUP	S98T002841	0	DSC-03	SOLID	<u>28.56</u>	<u>46.16</u>	N/A	Joules/g

Final page for worklist # 26792

[Signature] 10-26-98
Analyst Signature Date

[Signature] 10-29-98
Analyst Signature Date

Validated 10/29/98 [Signature]

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND102601 Mon Oct 26 10:18:55 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 417 TO 421.

N2, EXOTHERM DOWN
 Temp: 140.0 °C TIME: 0.0 min RATE: 10.0 °C/min
 KB MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 10:36:34 1998

Curve 1: DSC
 File Info: SAM102601 Mon Oct 26 11:55:28 1998
 Sample Weight: 10.110 mg
 S98T002840

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:45:49 1998

Curve 1: DSC
 File Info: SAM102602 Mon Oct 26 13:35:42 1998
 Sample Weight: 10.020 mg
 S98T002840 DUP

N2 10C/min
 Temp: 55.0 °C
 TIME: 8
 KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:48:04 1998

Curve 1: DSC
 File Info: SAM102603 Mon Oct 26 14:28:35 1998
 Sample Weight: 7.140 mg
 SS98T002841

N2 10C/min
 TEMPS: 38.8 °C
 TEND: 500.0 °C
 THERM: 0.0 min RATE: 10.0 C/min
 KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:50:15 1998

Curve 1: DSC
 File Info: SAM102604 Mon Oct 26 15:26:05 1998
 Sample Weight: 8.250 mg
 S98T002841 DUP

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:52:58 1998

N2 10C/min
 Temp: 35.0 °C
 Time: 550.0 s
 Rate: 0.0 min RATE: 10.0 C/min

LABCORE Data Entry Template for Worklist# 26793

Analyst: SLH Instrument: DSC0 3 Book # 12N14-BMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.03</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	2 SAMPLE	S98T002910	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	3 DUP	S98T002910	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	4 SAMPLE	S98T002911	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	5 DUP	S98T002911	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26793

Analyst Signature

Date

Analyst Signature

Date

Validated 10/29/98 88 MichaelKim Thomas 10-29-98

Data Entry Comments:

Trp ran on both S98T002910 + S98T002911 983
10/29/98Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND102702 Tue Oct 27 18:53:44 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 423 TO 427.

S. Hood Boatright

Curve 1: DSC
 File info: SAM102713 Tue Oct 27 20:28:09 1998
 Sample Weight: 13.020 mg
 S98T004910
 98102808

HNF-1663 REV. 0

Curve 1: DSC
 File info: SAM102714 Tue Oct 27 22:08:30 1998
 Sample Weight: 11.220 mg
 S98T002910 DUP

HNF-1663 REV. 0

Curve 1: DSC
 File info: SAM102801 Wed Oct 28 01:07:16 1998
 Sample Weight: 20.090 mg
 S98T001911 BUB SAM 08/10/98

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 01:19:30 1998

N2 10C/min
 Temp 25.8 °C
 Time 0.0 min RATE: 10.0 °C/min

HNF-1663 REV. 0

Curve 1: DSC

File info: SAM102802 Wed Oct 28 02:09:17 1998

Sample Weight: 20.070 mg

S98T001911 TTP-DOP 323 10/28/98

HNF-1663 REV. 0

LABCORE Data Entry Template for Worklist# 26794

Analyst: KRMInstrument: DSC0 3Book # 12N14BMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.17</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	2 SAMPLE	S98T002912	0	DSC-03	SOLID	<u>N/I</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	3 DUP	S98T002912	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	4 SAMPLE	S98T002913	0	DSC-03	SOLID	<u>N/I</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	5 DUP	S98T002913	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26794

Analyst Signature

10-27-98

Date

Analyst Signature

Date

Validated 11/3/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND102701 Tue Oct 27 06:26:21 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 422 TO 423

Curve 1: DSC
 File info: SAM102703 Tue Oct 27 07:36:58 1998
 Sample Weight: 16.710 mg
 S98T002912

Curve 1: DSC
 File Info: SAM02705 Tue Oct 27 09:50:31 1998
 Sample Weight: 10.690 mg
 S98T002912 DUP

N2 10C/min
 THERM: 200.0 g TIME: 0.0 min RATE: 10.0 C/min
 KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:57:21 1998

Curve 1: DSC
 File Info: SAM102706 Tue Oct 27 10:48:33 1998
 Sample Weight: 20.080 mg
 S98T002913

Curve 1: DSC
 File Info: SAM102707 Tue Oct 27 13:07:41 1998
 Sample Weight: 34.650 mg
 S98T002913 DUP

LABCORE Data Entry Template for Worklist# 26795

Analyst: KRM Instrument: DSC0 3 Book # 12N148

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>28.17x</u>	<u>N/A</u>	Joules/g
98000490	BY-105 (2)	2 SAMPLE	S98T002914	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000490	BY-105 (2)	3 DUP	S98T002914	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g
98000509	BY-105 (2)	4 SAMPLE	S98T002980	0	DSC-03	SOLID	<u>N/A</u>	<u>0</u>		Joules/g
98000509	BY-105 (2)	5 DUP	S98T002980	0	DSC-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26795

[Signature] 10-27-98
Analyst Signature Date

[Signature] 10-29-98
Analyst Signature Date

Validated 11/3/98 [Signature]

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND102701 Tue Oct 27 06:26:21 1998
 Sample Weight: 13.020 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 435 TO 439

Curve 1: DSC
 File Info: SAM102710 Tue Oct 27 16:04:18 1998
 Sample Weight: 15.250 mg
 S98T002980

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 16:28:49 1998

N2 10C/min
 Temp 300.0 °C
 TIME: 0.0 min RATE: 10.0 °C/min

Curve 1: DSC
 File Info: SAK102712 Tue Oct 27 17:29 1998
 Sample Weight: 13.760 mg
 S98T002980 DUP

Curve 1: DSC

File info: SAM102708 Tue Oct 27 14:00:26 1998

Sample Weight: 16.640 mg

S98T002914

N2 10C/min

TEMP: 35.0 °C

TIME: 555.0 s

0.0 min RATE: 10.0 C/min

KR MONTEITH
PERKIN-ELMER

7 Series Thermal Analysis System
Tue Oct 27 14:01:25 1998

HNF-1663 REV. 0

Curve 1: DSC
 File Info: SAK102709 Tue Oct 27 14:52:41 1998
 Sample Weight: 18.450 mg
 S98T002914 DUP

worklistrpt Version 2.1 05/15/95
10/23/98 10:40

Page: 1

LABCORE Data Entry Template for Worklist# 26796

Analyst: AK Instrument: DSC0 3 Boole # 12/14-13Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	28.45	28.22	N/A	Joules/g
98000509	BY-105 (2)	2 SAMPLE	S98T002981	0	DSC-03	SOLID	N/A	65.65	0	Joules/g
98000509	BY-105 (2)	3 DUP	S98T002981	0	DSC-03	SOLID	65.65	17.96	N/A	Joules/g
98000509	BY-105 (2)	4 TRIPLE	S98T002982	0	DSC-03	SOLID	65.65	16.01	0	Joules/g
98000509	BY-105 (2)	5 DUP	S98T002982	0	DSC-03	SOLID	0	0	N/A	Joules/g
98000509	BY-105 (2)	6 SAMPLE	S98T002983	0	DSC-03	SOLID	N/A	0		Joules/g
98000509	BY-105 (2)	7 DUP	S98T002983	0	DSC-03	SOLID	0	0	N/A	Joules/g

Final page for worklist # 26796

Analyst Signature Not Mon Date 10/22/98Analyst Signature Kim Thomas Date 10-29-98Validated 11/3/98 Dr. Machelor

Data Entry Comments:

High RPD due to sample inhomogeneities. Duplicate in good agreement with triplicate. No rerun 11/9/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND102601 Mon Oct 26 10:18:56 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 441 TO 442

N2, EXOTHERM DOWN

TEMP: 148.8 °C

OK 10/26/98
MONTIELLH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 02:07:24 1998

Curve 1: DSC
 File Info: SAM02607 Mon Oct 26 22:09:38 1998
 Sample Weight: 5.300 mg
 S98T002981

Curve 1: DSC
 File info: SAM102608 Mon Oct 26 23:03:42 1998
 Sample Weight: 9.400 mg
 S98T002981 DUP

Curve 1: DSC
 File Info: SAM102609 Tue Oct 27 00:08:46 1998
 Sample Weight: 9.150 mg
 S98T002981 TRIP

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 10:42:22 1998

Curve 1: DSC
 File Info: SAM102605 Mon Oct 26 19:52:13 1998
 Sample Weight: 11.780 mg
 S98T002982

RH KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 19:54:17 1998

N2 10C/min
 Temp: 38.8 °C
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: DSC
 File info: SAM102506 Mon Oct 26 20:45:43 1998
 Sample Weight: 27.000 mg
 S98T002982 DUP

Curve 1: DSC
 File Info: SAN102701 Tue Oct 27 00:59:15 1998
 Sample Weight: 17.990 mg
 S98T002983

Curve 1: DSC
 File Info: SAM102702 Tue Oct 27 02:45:31 1998
 Sample Weight: 12.050 mg
 S98T002983 DUP

LABCORE Completed Worklist Report for Worklist# 26852

Analyst: rwk

Instrument: DSC03

Book#: 12N14B

Method: LA-514-114

Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD		0	DSC-03	LIQUID	28.45	28.63*	100.633	% Recovery
2 SAMPLE	S98T003002	0	DSC-03	LIQUID	N/A	21.37		Joules/g
3 DUP	S98T003002	0	DSC-03	LIQUID	21.57	31.84	39.354	RPD
4 SAMPLE	S98T003003	0	DSC-03	LIQUID	N/A	53.02		Joules/g
5 DUP	S98T003003	0	DSC-03	LIQUID	53.02	60.64	13.408	RPD

Final page for worklist# 26852

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date _____

See next page for signatures *[Signature]* 2/3/99

LABCORE Data Entry Template for Worklist# 26852

Analyst: NK Instrument: DSC0 3 Book # 12/11/4-5Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.63*</u>	<u>N/A</u>	Joules/g
							<u>21.37</u>	<u>53.02</u>	<u>11/3/98</u>	Joules/g
98000510	BY-105 (2)	2 SAMPLE	S98T003002	0	DSC-03	LIQUID	<u>N/A</u>	<u>53.02</u>	<u>11/3/98</u>	Joules/g
							<u>21.37</u>	<u>31.84</u>	<u>11/3/98</u>	Joules/g
98000510	BY-105 (2)	3 DUP	S98T003002	0	DSC-03	LIQUID	<u>53.02</u>	<u>60.64</u>	<u>11/3/98</u>	Joules/g
							<u>53.02</u>	<u>53.02</u>	<u>11/3/98</u>	Joules/g
98000510	BY-105 (2)	4 SAMPLE	S98T003003	0	DSC-03	LIQUID	<u>N/A</u>	<u>53.02</u>	<u>11/3/98</u>	Joules/g
							<u>53.02</u>	<u>60.64</u>	<u>11/3/98</u>	Joules/g
98000510	BY-105 (2)	5 DUP	S98T003003	0	DSC-03	LIQUID	<u>21.37</u>	<u>31.84</u>	<u>11/3/98</u>	Joules/g
							<u>53.02</u>	<u>60.64</u>	<u>11/3/98</u>	Joules/g

Final page for worklist # 26852

Analyst Signature NK Date 10/30/98Analyst Signature Rim Thomas Date 11-5-98Validated 4/10/98 BTachebBTacheb

Data Entry Comments:

High RPD S98T003002 due to sample inhomogeneities. Rerun only at customer request. BT 4/3/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File Info: IND103601 Fri Oct 30 08:39:58 1998

Sample Weight: 13.020 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 451 TO 455

Curve 1: DSC
 File Info: SAM103011 Fri Oct 30 16:31:44 1998
 Sample Weight: 11.070 mg
 S98T003002

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 15:44:17 1998

Curve 1: DSC

File Info: SAM103012 Fri Oct 30 17:39:30 1998

Sample Weight: 13.650 mg

S98T003002 DUP

AW KING
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Nov 3 15:46:43 1998

Curve 1: DSC
 File Info: SAM103009 Fri Oct 30 14:47:02 1998
 Sample Weight: 12.880 mg
 S98T002003

02/13/98

RH KING
 PERKIN-ELMER
 / Series Thermal Analysis System
 Tue Nov 3 15:38:45 1998

Curve 1: DSC
 File info: SAM103010 Fri Oct 30 15:40:28 1998
 Sample Weight: 14.860 mg
 S9BT003003 DVP 11/16

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 15:40:48 1998

LABCORE Data Entry Template for Worklist# 26854

Analyst: SCH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D1

Worklist Comment: xBY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.19</u>	<u>N/A</u>	Joules/g
98000510	BY-105 (2)	2 SAMPLE	S98T003004	0	DSC-03	LIQUID	<u>N/A</u>	<u>8.25</u>		Joules/g
98000510	BY-105 (2)	3 DUP	S98T003004	0	DSC-03	LIQUID	<u>8.25</u>	<u>7.29</u>	<u>N/A</u>	Joules/g
98000510	BY-105 (2)	4 SAMPLE	S98T003005	0	DSC-03	LIQUID	<u>N/A</u>	<u>17.54</u>		Joules/g
98000510	BY-105 (2)	5 DUP	S98T003005	0	DSC-03	LIQUID	<u>17.54</u>	<u>23.52</u>	<u>N/A</u>	Joules/g

Final page for worklist # 26854

Sandra Hood Boatright
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Validated 11/12/98 [Signature]

Data Entry Comments:

Ran Trip on S98T 003004 11/10/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND102801 Wed Oct 28 16:37:14 1998
 Sample Weight: 11.670 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 451 TO 461.

N2, EXOTHERM DOWN
 140.0 °C TIME: 0.0 min RATE: 40.0 °/min
 SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 16:56:09 1998
SL Hood Boat right

Curve 1: DSC
 File info: SAM102901 Thu Oct 29 01:45:30 1998
 Sample Weight: 16.690 mg
 S98T003004

N2 10C/min
 THERMISTOR
 THERMISTOR

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Nov 2 14:54:32 1998

Curve 1: DSC
 File info: SAM102902 Thu Oct 29 02:37:41 1998
 Sample Weight: 16.000 mg
 S98T003004 DUP

N2 10C/min
 Temp: 35.0 °C
 Time: 00:00

TIME: 0.0 min DATE: 10.0 C/min

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Nov 2 14:59:27 1998

Curve 1: DSC

File Info: SAM102904 Thu Oct 29 04:45:11 1998

Sample Weight: 15.900 mg

S98T003005

N2 10C/min

TIME: 35.0 C

RATE: 10.0 C/min

Curve 1: DSC

File Info: SAM102905 Thu Oct 29 05:50:17 1998

Sample Weight: 19.370 mg

S98T003005 DUP

N2 10C/min

TEMP: 500.0 °C

TEMP: 50.0 °C

Temperature (°C)

SL HOOD BOATRRIGHT
PERKIN-ELMER7 Series Thermal Analysis System
Mon Nov 2 15:04:05 1998

LABCORE Data Entry Template for Worklist# 26916

Analyst: ppb Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: dsc-02, LIQUID, BY105 cjo

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000509	BY-105 (2)	1 SAMPLE	S98T002957	0	DSC-02	LIQUID	<u>N/A</u>	<u>9.69</u>		Joules/g Dry
98000509	BY-105 (2)	2 DUP	S98T002957	0	DSC-02	LIQUID	<u>9.69</u>	<u>30.87</u>	<u>N/A</u>	Joules/g Dry
98000509	BY-105 (2)	3 TRIPL	S98T002957	0	DSC-02	LIQUID	<u>9.69</u>	<u>28.40</u>	<u>N/A</u>	Joules/g Dry
98000509	BY-105 (2)	4 SAMPLE	S98T002958	0	DSC-02	LIQUID	<u>N/A</u>	<u>21.77</u>		Joules/g Dry
98000509	BY-105 (2)	5 DUP	S98T002958	0	DSC 02	LIQUID	<u>21.77</u>	<u>17.00</u>	<u>N/A</u>	Joules/g Dry
98000509	BY-105 (2)	6 SAMPLE	S98T002959	0	DSC-02	LIQUID	<u>N/A</u>	<u>23.97</u>		Joules/g Dry
98000509	BY-105 (2)	7 DUP	S98T002959	0	DSC-02	LIQUID	<u>23.97</u>	<u>23.44</u>	<u>N/A</u>	Joules/g Dry

Final page for worklist # 26916

ppb 11/12/98
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26917

Analyst: PPB Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, SOLID, BY-105 cjo

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000436	BY-105 (2)	1 SAMPLE	S98T002589	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000436	BY-105 (2)	2 DUP	S98T002589	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000436	BY-105 (2)	3 SAMPLE	S98T002590	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000436	BY-105 (2)	4 DUP	S98T002590	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000437	BY-105 (2)	5 SAMPLE	S98T002840	0	DSC-02	SOLID	N/A	42.28		Joules/g Dry
98000437	BY-105 (2)	6 DUP	S98T002840	0	DSC-02	SOLID	42.28	33.80	N/A	Joules/g Dry
98000490	BY-105 (2)	7 SAMPLE	S98T002910	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	8 DUP	S98T002910	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000490	BY-105 (2)	9 SAMPLE	S98T002911	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	10 DUP	S98T002911	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000490	BY-105 (2)	11 SAMPLE	S98T002912	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	12 DUP	S98T002912	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry

Final page for worklist # 26917

PPB 11/13/98
Analyst Signature Date

Rim Thomas 11-17-98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26918

Analyst: ppb Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, SOLID, BY-105 cjo

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000490	BY-105 (2)	1 SAMPLE	S98T002913	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	2 DUP	S98T002913	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000490	BY-105 (2)	3 SAMPLE	S98T002914	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	4 DUP	S98T002914	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000509	BY-105 (2)	5 SAMPLE	S98T002980	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000509	BY-105 (2)	6 DUP	S98T002980	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000509	BY-105 (2)	7 SAMPLE	S98T002981	0	DSC-02	SOLID	N/A	109.3		Joules/g Dry
98000509	BY-105 (2)	8 DUP	S98T002981	0	DSC-02	SOLID	109.3	24.90	N/A	Joules/g Dry
98000509	BY-105 (2)	9 TRIPL	S98T002981	0	DSC-02	SOLID	109.3	26.65	N/A	Joules/g Dry
98000509	BY-105 (2)	10 SAMPLE	S98T002982	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000509	BY-105 (2)	11 DUP	S98T002982	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry
98000509	BY-105 (2)	12 SAMPLE	S98T002983	0	DSC-02	SOLID	N/A	0		Joules/g Dry
98000509	BY-105 (2)	13 DUP	S98T002983	0	DSC-02	SOLID	0	0	N/A	Joules/g Dry

Final page for worklist # 26918

PPB
Analyst Signature Date 11/13/98

Kim Thomas
Analyst Signature Date 11-17-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27166

Analyst: JS Instrument: DSC0 3 Book # 12N14B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY105 (2), DSC-03 Run under nitrogen. skm

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				DSC-03	LIQUID	<u>28.45</u>	<u>27.87</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	2	SAMPLE	S98T002811	0		DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g
98000437	BY-105 (2)	3	DUP	S98T002811	0		DSC-03	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 27166

JS JS 11/12/98
Analyst Signature Date

Kim Thomas 11-17-98
Analyst Signature Date

Validated 11/18/98 JS Bachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File Info: IND111201 Thu Nov 12 19:14:18 1998
 Sample Weight: 6.530 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 466 TO 468.

N2, EXOTHERM DOWN
 TEMP: 140.0 °C
 TIME: 180.0 s

Temperature (°C)

0.0 min RATE: 10.0 °C/min

JD SPELLMAN
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Nov 12 19:15:05 1998

Curve 1: DSC
File Info: SAM11201 Thu Nov 12 20:29:47 1998
Sample Weight: 15.320 mg
S98T002811

N2 10C/min
TEMP: 50.0 °C
TIME: 0.0 min RATE: 10.0 C/min
JD SPELLMAN
PERKIN-ELMER
7 Series Thermal Analysis System
Thu Nov 12 20:34:25 1998

Curve 1: DSC
File Info: SAM11202 Thu Nov 12 21:24:19 1998
Sample Weight: 13.590 mg
S98T002811DUP

JD SPELLMAN
PERKIN-ELMER
7 Series Thermal Analysis System
Thu Nov 12 21:36:24 1998

N2 10C/min
TEMP: 35.0 °C
TIME: 500.0 s
RATE: 0.0 min RATE: 10.0 C/min

LABCORE Data Entry Template for Worklist# 27173

Analyst: ppb Instrument: DSC0 Book # N/A

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, LIQUID, BY-105 cjo

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
98000436	BY-105 (2)	1 SAMPLE	S98T002392 0		DSC-02	LIQUID	N/A	0		Joules/g Dry
98000436	BY-105 (2)	2 DUP	S98T002392 0		DSC-02	LIQUID	0	0	N/A	Joules/g Dry
98000437	BY-105 (2)	3 SAMPLE	S98T002806 0		DSC-02	LIQUID	N/A	28.20		Joules/g Dry
98000437	BY-105 (2)	4 DUP	S98T002806 0		DSC-02	LIQUID	28.20	24.30	N/A	Joules/g Dry
98000437	BY-105 (2)	5 SAMPLE	S98T002807 0		DSC-02	LIQUID	N/A	25.51		Joules/g Dry
98000437	BY-105 (2)	6 DUP	S98T002807 0		DSC-02	LIQUID	25.51	26.75	N/A	Joules/g Dry
98000437	BY-105 (2)	7 SAMPLE	S98T002808 0		DSC-02	LIQUID	N/A	20.48		Joules/g Dry
98000437	BY-105 (2)	8 DUP	S98T002808 0		DSC-02	LIQUID	20.48	23.08	N/A	Joules/g Dry
98000437	BY-105 (2)	9 SAMPLE	S98T002813 0		DSC-02	LIQUID	N/A	5.52		Joules/g Dry
98000437	BY-105 (2)	10 DUP	S98T002813 0		DSC-02	LIQUID	5.52	3.01	N/A	Joules/g Dry
98000490	BY-105 (2)	11 SAMPLE	S98T002938 0		DSC-02	LIQUID	N/A	0		Joules/g Dry
98000490	BY-105 (2)	12 DUP	S98T002938 0		DSC-02	LIQUID	0	0	N/A	Joules/g Dry
98000509	BY-105 (2)	13 SAMPLE	S98T002955 0		DSC-02	LIQUID	N/A	58.69		Joules/g Dry
98000509	BY-105 (2)	14 DUP	S98T002955 0		DSC-02	LIQUID	58.69	71.64	N/A	Joules/g Dry
98000509	BY-105 (2)	15 SAMPLE	S98T002956 0		DSC-02	LIQUID	N/A	62.22		Joules/g Dry
98000509	BY-105 (2)	16 DUP	S98T002956 0		DSC-02	LIQUID	62.22	61.51	N/A	Joules/g Dry
98000510	BY-105 (2)	17 SAMPLE	S98T003002 0		DSC-02	LIQUID	N/A	44.96		Joules/g Dry
98000510	BY-105 (2)	18 DUP	S98T003002 0		DSC-02	LIQUID	44.96	66.98	N/A	Joules/g Dry

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27173

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000510	BY-105 (2)	19 SAMPLE	S98T003003	0	DSC-02	LIQUID	N/A	109.39		Joules/g Dry
98000510	BY-105 (2)	20 DUP	S98T003003	0	DSC-02	LIQUID	109.39	125.11	N/A	Joules/g Dry
98000437	BY-105 (2)	21 SAMPLE	S98T002809	0	DSC-02	LIQUID	N/A	34.36		Joules/g Dry
98000437	BY-105 (2)	22 DUP	S98T002809	0	DSC-02	LIQUID	34.36	58.04	N/A	Joules/g Dry
98000437	BY-105 (2)	23 SAMPLE	S98T002810	0	DSC-02	LIQUID	N/A	46.67		Joules/g Dry
98000437	BY-105 (2)	24 DUP	S98T002810	0	DSC-02	LIQUID	46.67	53.46	N/A	Joules/g Dry

Final page for worklist # 27173

Analyst Signature

Date
11/18/98

Analyst Signature
11/18/98
Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27174

Analyst: APB Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, SOLID, BY-105, cjo

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000436	BY-105 (2)	1 SAMPLE	S98T002591	0	DSC-02	SOLID	N/A	7.53		Joules/g Dry
98000436	BY-105 (2)	2 DUP	S98T002591	0	DSC-02	SOLID	7.53	6.28	N/A	Joules/g Dry
98000436	BY-105 (2)	3 SAMPLE	S98T002592	0	DSC-02	SOLID	N/A	1.63		Joules/g Dry
98000436	BY-105 (2)	4 DUP	S98T002592	0	DSC-02	SOLID	1.63	1.82	N/A	Joules/g Dry
98000436	BY-105 (2)	5 SAMPLE	S98T002593	0	DSC-02	SOLID	N/A	344.82		Joules/g Dry
98000436	BY-105 (2)	6 DUP	S98T002593	0	DSC-02	SOLID	344.82	248.68	N/A	Joules/g Dry
98000436	BY-105 (2)	7 SAMPLE	S98T002594	0	DSC-02	SOLID	N/A	252.86		Joules/g Dry
98000436	BY-105 (2)	8 DUP	S98T002594	0	DSC-02	SOLID	252.86	337.8	N/A	Joules/g Dry
98000436	BY-105 (2)	9 SAMPLE	S98T002711	0	DSC-02	SOLID	N/A	29.57		Joules/g Dry
98000436	BY-105 (2)	10 DUP	S98T002711	0	DSC-02	SOLID	29.57	25.24	N/A	Joules/g Dry
98000490	BY-105 (2)	11 SAMPLE	S98T002737	0	DSC-02	SOLID	N/A	20.78		Joules/g Dry
98000490	BY-105 (2)	12 DUP	S98T002737	0	DSC-02	SOLID	20.78	23.83	N/A	Joules/g Dry

Final page for worklist # 27174

APB 11/18/98
Analyst Signature Date

APB 11/18/98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27230

Analyst: ppb Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, BY105, JDS

GROUP	PROJECT	S TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000437	BY-105 (2)	1 SAMPLE	S98T002811	0		DSC-02	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g Dry
98000437	BY-105 (2)	2 DUP	S98T002811	0		DSC-02	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g Dry
98000437	BY-105 (2)	3 SAMPLE	S98T002812	0		DSC-02	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g Dry
98000437	BY-105 (2)	4 DUP	S98T002812	0		DSC-02	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g Dry
98000510	BY-105 (2)	5 SAMPLE	S98T003004	0		DSC-02	LIQUID	<u>N/A</u>	<u>17.37</u>		Joules/g Dry
98000510	BY-105 (2)	6 DUP	S98T003004	0		DSC-02	LIQUID	<u>17.37</u>	<u>15.35</u>	<u>N/A</u>	Joules/g Dry
98000510	BY-105 (2)	7 SAMPLE	S98T003005	0		DSC-02	LIQUID	<u>N/A</u>	<u>36.63</u>		Joules/g Dry
98000510	BY-105 (2)	8 DUP	S98T003005	0		DSC-02	LIQUID	<u>36.63</u>	<u>49.12</u>	<u>N/A</u>	Joules/g Dry

Final page for worklist # 27230

[Signature] 11/18/98
Analyst Signature Date

[Signature] 11/18/98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27231

Analyst: ppb Instrument: DSC0 Book # N/A

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: DSC-02, SOLID, BY105, JDS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000437	BY-105 (2)	1 SAMPLE	S98T002841	0	DSC-02	SOLID	<u>N/A</u>	<u>36.09</u>		Joules/g Dry
98000437	BY-105 (2)	2 DUP	S98T002841	0	DSC-02	SOLID	<u>36.09</u>	<u>58.73</u>	<u>N/A</u>	Joules/g Dry

Final page for worklist # 27231

JAL 11/18/98
Analyst Signature Date

Rachel 11/18/98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Completed Worklist Report for Worklist# 27633

Analyst: slh

Instrument: DSC03

Book#: 12N14B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 DSC-03 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		DSC-03	LIQUID	28.45	27.25	95.782 % Recovery	
2 SAMPLE	S98T003383	0	DSC-03	LIQUID	N/A	0	Joules/g	
3 DUP	S98T003383	0	DSC-03	LIQUID	0	0	0.000 EPD	

Final page for worklist# 27633

Analyst Signature _____ Date _____

Mary Fran 12-29-98
Analyst Signature Date

B. Pacheco 12/30/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27633

Analyst: SCH Instrument: DSC0 3 Book # 12N14-B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 DSC-03 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>27.25</u>	<u>N/A</u>	Joules/g
98000437	BY-105 (2)	2 SAMPLE	S98T003383	0	DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g
98000437	BY-105 (2)	3 DUP	S98T003383	0	DSC-03	LIQUID	<u>0</u>	<u>0</u>	<u>N/A</u>	Joules/g

Final page for worklist # 27633

Andrea Wood Boatright
Analyst Signature Date
12-20-98

Mary Tracy
Analyst Signature Date
12-29-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND122002 Sun Dec 20 07: 44: 13 1998

Sample Weight: 19.460 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 476 TO 478

SL HOOD BOATRACHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 07: 59: 54 1998

Signature of Howard Boatright

Curve 1: DSC
 File info: SAM122005 Sun Dec 20 09: 03: 16 1998
 Sample Weight: 13.370 mg
 S98T003983

SL HOOD BOATRHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 09: 53: 15 1998

Curve 1: DSC
 File info: SAM122006 Sun Dec 20 10:44:46 1998
 Sample Weight: 14.600 mg
 S98T003383 DUP

N2 10C/min
 TIME: 38.8 g
 500.0 g

SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 11:08:46 1998

LABCORE Completed Worklist Report for Worklist# 27634

Analyst: rwk

Instrument: DSC03

Book#: 12N14B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 DSC-03 RLC

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0	DSC-03	SOLID	28.45	27.50	96.661 % Recovery	
2	SAMPLE	S98T003391	0	DSC-03	SOLID	N/A	11.54	Joules/g	
3	DUP	S98T003391	0	DSC-03	SOLID	11.54	21.24	59.182 RPD	
4	SAMPLE	S98T003396	0	DSC-03	SOLID	N/A	125.6	Joules/g	
5	DUP	S98T003396	0	DSC-03	SOLID	125.6	130.3	3.673 RPD	

Final page for worklist# 27634

Analyst Signature _____ Date _____

Mary J. Mary 12-29-98
Analyst Signature Date

Q. Bachelar 12/30/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27634

Analyst: JK Instrument: DSC0 3 Book # 12/14-13Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 DSC-03 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>27.50</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T003391	0	DSC-03	SOLID	<u>N/A</u>	<u>11.54</u>		Joules/g
98000436	BY-105 (2)	3 DUP	S98T003391	0	DSC-03	SOLID	<u>11.54</u>	<u>21.24</u>	<u>N/A</u>	Joules/g
98000436	BY-105 (2)	4 SAMPLE	S98T003396	0	DSC-03	SOLID	<u>N/A</u>	<u>125.6</u>		Joules/g
98000436	BY-105 (2)	5 DUP	S98T003396	0	DSC-03	SOLID	<u>125.6</u>	<u>130.3</u>	<u>N/A</u>	Joules/g

Final page for worklist # 27634

Analyst Signature JK Date 12/20/98Analyst Signature Mary Fran Date 12-29-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File info: IND121901 Sat Dec 19 18:45:45 1998
 Sample Weight: 19.460 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 481 TO 485.

Curve 1: DSC

File info: SAM121901 Sat Dec 19 19: 41: 46 1998

Sample Weight: 16.490 mg

S98T003391

N2 10C/min
TEMP 50.0 °C
TIME 0.0

MIN RATE: 10.0 C/min

Temperature (°C)

HW KINS
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Dec 29 08:37:07 1998

Curve 1: DSC

File Info: SAM121903 Sat Dec 19 22:11:08 1998

Sample Weight: 15.580 mg

S98T003391 TRIP-DIP *SB* 12/20/98

N2 10C/min

TEMP: 285.8 °C

TIME: 0.0 min

RATE: 10.0 C/min

Temperature (°C)

RM KINS

PERKIN-ELMER

7 Series Thermal Analysis System

Tue Dec 29 08:41:32 1998

Curve 1: DSC
 File Info: SAM122002 Sun Dec 20 02:20:48 1998
 Sample Weight: 16.330 mg
 S98T003396

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Dec 29 08:44:57 1998

N2 10C/min
 Temp: 50.0 °C
 Time: 558.0 s
 Rate: 10.0 C/min

Curve 1: DSC

File Info: SAM122003 Sun Dec 20 03:14:58 1998

Sample Weight: 30.910 mg

S98T003396 DUP

HNF-1663 REV. 0

N2 10C/min

TEMP: 500.0 g

TIME: 0.0 min RATE: 10.0 C/min

Temperature (°C)

RM KING
PERKIN-ELMER

7 Series Thermal Analysis System
Tue Dec 29 08:46:52 1998

LABCORE Completed Worklist Report for Worklist# 27643

Analyst: rwk

Instrument: DSC03

Book#: 12N14B

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD	0		DSC-03	SOLID	28.45	27.48 *	96.531 % Recovery	
2	SAMPLE	S98T003386	0	DSC-03	SOLID	N/A	43.74	Joules/g	
3	DUP	S98T003386	0	DSC-03	SOLID	43.74	38.04	13.940 RPD	

Final page for worklist# 27643

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

worklistrpt Version 2.1 05/15/95
12/18/98 10:12

Page: 1

LABCORE Data Entry Template for Worklist# 27643

Analyst: RL Instrument: DSC0 3 Book # 12/14-13Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	SOLID	<u>28.45</u>	<u>27.48</u>	N/A	Joules/g
98000490	BY-105 (2)	2 SAMPLE	S98T003386	0	DSC-03	SOLID	N/A	<u>43.74</u>		Joules/g
98000490	BY-105 (2)	3 DUP	S98T003386	0	DSC-03	SOLID	<u>43.74</u>	<u>38.64</u>	N/A	Joules/g

Final page for worklist # 27643

Analyst Signature: Not Rnd Date: 12/20/98Analyst Signature: Mary Fran Date: 12-29-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: DSC

File info: IND121901 Sat Dec 19 18:45:45 1998

Sample Weight: 19.460 mg

STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 488 TO 490

N2, EXOTHERM DOWN

TEMP: 168.8 °C

TIME: 8

0.0 min RATE: 10.0 C/min

Temperature (°C)

RH KING
PERKIN-ELMER

7 Series Thermal Analysis System
Sat Dec 19 22:52:31 1998

Curve 1: DSC
 File Info: SAH121904 Sat Dec 19 23:01:35 1998
 Sample Weight: 22.310 mg
 9987003386

N2 10C/min
 TEMPERATURE 20.0 °C
 TIME: 0.0 min RATE: 10.0 C/min
 RM KINS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Dec 29 08:27:36 1998

Curve 1: DSC
 File info: SAM122001 Sun Dec 20 01:10:35 1998
 Sample Weight: 21.050 mg
 S98T003386 DUP

490

N2 10C/min
 TIME: 25.8 g
 RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Dec 29 08:30:42 1998

LABCORE Completed Worklist Report for Worklist# 27765

Analyst: ppb

Instrument: DSC0

Book#: NA

Method: LA-514-1134 ^{01/15/99} Rev/Mod D-1

Worklist Comment: BY-105, DSC-02, solid cjo

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 SAMPLE	S98T003386 0	DSC-02	SOLID	N/A	45.31		Joules/g Dry
2 DUP	S98T003386 0	DSC-02	SOLID	45.31	39.40	13.953	RPD
3 SAMPLE	S98T003391 0	DSC-02	SOLID	N/A	11.82		Joules/g Dry
4 DUP	S98T003391 0	DSC-02	SOLID	11.82	21.76	59.202	RPD
5 SAMPLE	S98T003396 0	DSC-02	SOLID	N/A	181.03		Joules/g Dry
6 DUP	S98T003396 0	DSC-02	SOLID	181.03	187.81	3.676	RPD

Final page for worklist# 27765

B. Micheloz 1/5/99
Analyst Signature Date

C. Opiner 1/4/99
Analyst Signature Date

_____
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27811

Analyst: ppb Instrument: DSC0 Book # NA

Method: LA-514-113 Rev/Mod D-1

Worklist Comment: BY-105 DRY DSC LIQUID MF

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
98000437	BY-105 (2)	1 SAMPLE	S98T003383	0	DSC-02	LIQUID	N/A	<u>⊕</u>		Joules/g Dry
98000437	BY-105 (2)	2 DUP	S98T003383	0	DSC-02	LIQUID	<u>⊕</u>	<u>⊕</u>	N/A	Joules/g Dry

Final page for worklist # 27811

B. Bachelor 1/7/99
Analyst Signature Date

Mary Irene 1-7-99
Analyst Signature Date

Data Entry Comments:

LABCORE Completed Worklist Report for Worklist# 28062

Analyst: rwk

Instrument: DSC03

Book#: 12N14B

Method: LA-514-114 Rev/Mod _____

Worklist Comment: BY105 FIELD BLANK FOR DSC-03 RTS

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0		DSC-03	LIQUID	28.45	28.56	100.387	% Recovery
2	SAMPLE	S98T002391	0		DSC-03	LIQUID	N/A	0		Joules/g
3	DUP	S98T002391	0		DSC-03	LIQUID	0	0	0.000	RED

Final page for worklist# 28062

_____
Analyst Signature

_____
Date

C. J. Opium 1/29/99
_____
Analyst Signature Date

_____
Reviewer Signature

_____
Date

worklistrpt Version 2.1 05/15/95
01/27/99 14:05

Page: 1

LABCORE Data Entry Template for Worklist# 28062

Analyst: ML Instrument: DSC0 0? Book # 12114-BMethod: LA-514-114 Rev/Mod D-2

Worklist Comment: BY105 FIELD BLANK FOR DSC-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			DSC-03	LIQUID	<u>28.45</u>	<u>28.56*</u>	N/A	Joules/g
98000436	BY-105 (2)	2 SAMPLE	S98T002391	0	DSC-03	LIQUID	<u>N/A</u>	<u>0</u>		Joules/g
98000436	BY-105 (2)	3 DUP	S98T002391	0	DSC-03	LIQUID	<u>0</u>	<u>0</u>	N/A	Joules/g

Final page for worklist # 28062

Kol. King 1/27/99
 Analyst Signature Date

C. J. O'Brien 1/29/99
 Analyst Signature Date
Validated 2/1/99 ML

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

Curve 1: DSC
 File info: IND012701 Wed Jan 27 08:39:37 1999
 Sample Weight: 19.460 mg
 STD 12N14-B

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 495 TO 497

N2, EXOTHERM DOWN
 TEMPE 160.0 °C TIME: 0.0 min RATE: 10.0 °/min
 LAD 7/2/99
 SL-HOOD-BD-ATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Jan 27 22:30:45 1999

Curve 1: DSC

File info: SAK012714 Wed Jan 27 22:11:49 1999

Sample Weight: 12.710 mg

S99T002391

HNF-1663 REV. 0

N2 100/min TIME: 5:08.8
TEMP: 508.8 °C
3.0 min RATE: 10.0 °C/min
RW KING PERKIN-ELMER
7 Series Thermal Analysis System
Wed Jan 27 22:19:42 1999

Curve 1: DSC
 File info: SAM012715 Wed Jan 27 23:11:20 1999
 Sample Weight: 15.910 mg
 S99T002991 DUP

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Jan 27 23:38:07 1999

N2 10C/min
 35.0 g
 0.0 min RATE: 10.0 C/min
 TIME: 555.8 s

LABCORE Completed Worklist Report for Worklist# 28888

Analyst: ppb

Instrument: DSC0

Book#: NA

Method: LA-514-113 Rev/Mod D-2

Worklist Comment: BY 105 DSC-O2 MF

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 SAMPLE	S98T002391	0	DSC-02	LIQUID	N/A	0		Joules/g Dry
2 DUP	S98T002391	0	DSC-02	LIQUID	0	0	0.000 RPD	

Final page for worklist# 28888

Analyst Signature _____ Date _____

Mary Tracy 3-19-99
Analyst Signature Date

P. Michelson 3/22/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26765

Analyst: PK Instrument: TGA0 3 Book # 11728-4

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	59.4	58.80	N/A	%
98000436	BY-105 (2)	2 SAMPLE	S98T002392 0		TGA-03	LIQUID	N/A	97.08		%
98000436	BY-105 (2)	5 DUP	S98T002392 0		TGA-03	LIQUID	97.08	96.92	N/A	%
98000437	BY-105 (2)	4 SAMPLE	S98T002806 0		TGA-03	LIQUID	N/A	52.76		%
98000437	BY-105 (2)	5 DUP	S98T002806 0		TGA-03	LIQUID	52.76	51.92	N/A	%

Final page for worklist # 26765

Not Kim 10/30/98
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Validated 11/11/98 BJachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER103001 Fri Oct 30 08:57:53 1998

Sample Weight: 19.574 mg

117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 500 TO 502.

500

10C/MIN N2

TEMP: 25.0 °C TIME: 0.0 MIN RATE: 10.0 °C/MIN

Temperature (°C)

RM KING

TERMINAL-ELIEN

7 Series Thermal Analysis System

Fri Oct 30 08:59:10 1998

HNF-1663 REV. 0

Curve 1: TGA

File Info: SAM03007 Fri Oct 30 12:38:19 1998

Sample Weight: 15.264 mg

5987002806

10C/MIN N2

TEMP: 35.0 °C
TIME: 000.0 s

6.0 min RATE: 10.0 °C/min

Temperature (°C)

BN KING
FURNACE
7 BRUIERBruiser Thermal Analysis System
Fri Oct 30 12:45:54 1998

Curve 1: TGA
 File info: SAM103006 Fri Oct 30 13:44:38 1998
 Sample Weight: 15.996 mg
 S92T002806 DUP

worklistrpt Version 2.1 05/15/95
10/23/98 09:58

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26766

Analyst: JEH Instrument: TGA0 3 Book # 117 N8-AMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.71</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002807	0	TGA-03	LIQUID	<u>N/A</u>	<u>53.04</u>		%
98000437	BY-105 (2)	3 DUP	S98T002807	0	TGA-03	LIQUID	<u>53.04</u>	<u>53.34</u>	<u>N/A</u>	%
98000437	BY-105 (2)	4 SAMPLE	S98T002808	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.86</u>		%
98000437	BY-105 (2)	5 DUP	S98T002808	0	TGA-03	LIQUID	<u>52.86</u>	<u>52.84</u>	<u>N/A</u>	%

Final page for worklist # 26766

Analyst Signature Imelda Wood Boatright
Date 10-30-98Analyst Signature Kim Thomas
Date 11-5-98Validated 11/11/98 AS Rachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

Curve 1: T8A

File Info: TER102901 Thu Oct 29 10:35:41 1998

Sample Weight: 14.549 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 504 TO 502.

Curve 1: TGA
 File Info: SAK103001 Fri Oct 30 01:50:34 1998
 Sample Weight: 16.141 mg
 9877002807

Curve 1: TGA
 File Info: SAM103002 Fri Oct 30 02:56:39 1998
 Sample Weight: 16.425 mg
 998T002807 DUP

SL HOOD BOATRUGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 03:36:04 1998

Curve 1: TGA

File Info: SAM103003 Fri Oct 30 04:25:46 1998

Sample Weight: 14.561 mg

8987002808

507

Curve 1: TSA
 File info: SAM103004 Fri Oct 30 05:31:14 1998
 Sample Weight: 15.490 mg
 S98T002808 DUP

SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 07:59:35 1998

100/MIN N2
 35.8 g
 500.0 g
 0.0 min RATE: 10.0 C/min
 TIME: 5

LABCORE Data Entry Template for Worklist# 26767

Analyst: DIS Instrument: TGA0 3 Book # 117NR-A

Method: LA-514-114 Rev/Mod 0.1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.71</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002809	0	TGA-03	LIQUID	<u>N/A</u>	<u>53.99</u>		%
98000437	BY-105 (2)	3 DUP	S98T002809	0	TGA-03	LIQUID	<u>53.99</u>	<u>53.46</u>	<u>N/A</u>	%
98000437	BY-105 (2)	4 SAMPLE	S98T002810	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.72</u>		%
98000437	BY-105 (2)	5 DUP	S98T002810	0	TGA-03	LIQUID	<u>52.72</u>	<u>52.78</u>	<u>N/A</u>	%

Final page for worklist # 26767

Jeff Holbrook 10-29-98
Analyst Signature Date

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/11/98 JB Maclebo

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File Info: TER102901 Thu Oct 29 10:35:41 1998
 Sample Weight: 14.549 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 510 TO 514.

510

10C/MIN N2
 TGA: 300.0 g THERM: 0.0 min RATE: 10.0 C/min
 JI SOLBRACK *JS*
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 21:52:26 1998

4NF-1663 REV. 0

Curve 1: TGA
 File info: SAM402809 Thu Oct 28 13:52:17 1998
 Sample Weight: 14.083 mg
 S98T002809

JT SOLBRACK
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 14:05:28 1998

Curve 1: TGA

File Info: SAM02910 Thu Oct 29 15:03:16 1998

Sample Weight: 14.099 mg

S98T0028090LP

J1 SOLBRACK
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 15:06:25 1998

Curve 1: TGA

File info: SAM102911 Thu Oct 29 16:11:57 1998

Sample Weight: 13.903 mg

5987002810

JT SOLBRACK
PERKIN-ELMER
7 Series Thermal Analysis System
Thu Oct 29 16:26:25 1998

Temperature (°C)

10.0 C/min

10C/MIN N2

TIME: 00:08

0.0 MIN RATE: 10.0 C/min

Curve 1: TGA
 File info: qsav1
 Thu Oct 29 19:15:45 1998
 Sample Weight: 15.878 mg
 S987002810 DUP

JT SOLBRACK
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 19:50:31 1998

LABCORE Data Entry Template for Worklist# 26768

Analyst: SCH Instrument: TGA0 3 Book # 117N8-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.98</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002811 0		TGA-03	LIQUID	<u>N/A</u>	<u>48.69</u>		%
98000437	BY-105 (2)	3 DUP	S98T002811 0		TGA-03	LIQUID	<u>48.69</u>	<u>52.44</u>	<u>N/A</u>	%
98000437	BY-105 (2)	4 SAMPLE	S98T002812 0		TGA-03	LIQUID	<u>N/A</u>	<u>52.20</u>		%
98000437	BY-105 (2)	5 DUP	S98T002812 0		TGA-03	LIQUID	<u>52.20</u>	<u>52.10</u>	<u>N/A</u>	%

Final page for worklist # 26768

Sandra H. Hood Boatright
Analyst Signature Date 11-3-98

11/3/98
Analyst Signature Date

Validated 11/2/98 S. Machelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File Info: TER110301 Tue Nov 3 07:01:57 1998
 Sample Weight: 19.677 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 516 TO 520.

Curve 1: TGA
 File info: SAM110301 Tue Nov 3 08:07:03 1998
 Sample Weight: 19.677 mg
 59819029500P Q446 11-2-48
 59819029500P

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 08:33:14 1998

10C/MIN N2
 TIME: 35.8 S
 RATE: 10.0 G/MIN

Curve 1: TGA
 File info: SAM110302 Tue Nov 3 09:23:02 1998
 Sample Weight: 16.142 mg
 S98T002811 DUP

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 09:49:55 1998

Curve 1: T6A
 File info: SAM10303 Tue Nov 3 10:47:18 1998
 Sample Weight: 16.100 mg
 98070005584P-2443 11-3-98
 2987000001

SL HODD BOATRIGHT
 PEKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 11:03:15 1998

10C/MIN N2
 0.0 MIN RATE: 10.0 C/min
 TIME: 00:00:00

Curve 1: TGA
 File info: SAM110304 Tue Nov 3 11:53:53 1998
 Sample Weight: 13.942 mg
 89870025500P- SL4B 11-3-98
 59870025500P- SL4B 11-3-98

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 12:45:42 1998

10C/MIN N2
 THERM 300.0 g
 TIME: 0.0 min RATE: 10.0 C/min

LABCORE Data Entry Template for Worklist# 26769

11-4-98

Analyst: RDA SLH Instrument: TGA0 3 Book # 117N8A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>57.4</u>	<u>58.91</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002813	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.14</u>		%
98000437	BY-105 (2)	3 DUP	S98T002813	0	TGA-03	LIQUID	<u>52.14</u>	<u>52.28</u>	<u>N/A</u>	%
98000490	BY-105 (2)	4 SAMPLE	S98T002938	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.02</u>		%
98000490	BY-105 (2)	5 DUP	S98T002938	0	TGA-03	LIQUID	<u>52.02</u>	<u>48.06</u>	<u>N/A</u>	%

Final page for worklist # 26769

RDA 11/2/98
Analyst Signature Date

SLH 11/9/98
Analyst Signature Date

Sandra L. Hord Boatright

11-4-98

Validated 11/9/98 SLH

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: T6A
 File Info: TER110401 Wed Nov 4 07:29:40 1998
 Sample Weight: 24.035 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 222 TO 226

10C/MIN N2
 TEND: 350.0 °C
 TIME: 0.0 MIN RATE: 10.0 °C/MIN

SL HOOD-80ATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Nov 4 12:41:32 1998

W. Pacheco

Curve 1: TGA
 File info: SAM110401 Wed Nov 4 13:32:01 1998
 Sample Weight: 18.428 mg
 S98T002B13

SL HOOD SDATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Nov 4 13:44:04 1998

10C/MIN N2
 0.0 min RATE: 10.0 C/min
 TIME: 8
 266.8 g

Curve 1: TSA
 File info: SAM110402 Wed Nov 4 14:41:38 1998
 Sample Weight: 16.425 mg
 S98T002813 DUP

Curve 1: TGA
 File info: SAM110403 Wed Nov 4 16:57:46 1998
 Sample Weight: 16.189 mg
 996T002938

10C/MIN N2
 TIME: 550.8 S
 TEMPE: 550.8 S
 SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Nov 4 16:23:08 1998

Curve 1: TGA
 File Info: SAM110404 Wed Nov 4 17:12:38 1998
 Sample Weight: 16.189 mg
 S98T002938 DUP

LABCORE Data Entry Template for Worklist# 26770

Analyst: RDM Instrument: TGA0 3 Book # 177N01A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				TGA-03	LIQUID	<u>59.4</u>	<u>58.80x</u>	<u>N/A</u>	%
98000509	BY-105 (2)	2	SAMPLE	S98T002955	0		TGA-03	LIQUID	<u>N/A</u>	<u>53.16</u>		%
98000509	BY-105 (2)	3	DUP	S98T002955	0		TGA-03	LIQUID	<u>53.16</u>	<u>52.59</u>	<u>N/A</u>	%
98000509	BY-105 (2)	4	SAMPLE	S98T002956	0		TGA-03	LIQUID	<u>N/A</u>	<u>52.80</u>		%
98000509	BY-105 (2)	5	DUP	S98T002956	0		TGA-03	LIQUID	<u>52.80</u>	<u>52.38</u>	<u>N/A</u>	%

Final page for worklist # 26770

RDM 10/31/98
Analyst Signature Date

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/16/98 Rachel

Data Entry Comments:

Duplicate for S98T002955 was only run to 300°C. No weight loss observed
above 300°C during sample analysis. Rerun only at customer request. 11/16/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER103001 Fri Oct 30 08:57:53 1998

Sample Weight: 19.574 mg

117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 528 TO 532

RM KING *B. Nachter*
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Nov 3 15:30:44 1998

Curve 1: TGA

File info: SAM103101 Sat Oct 31 00: 44: 14 1998

Sample Weight: 13.633 mg

S98T002955

Curve 1: TGA
 File info: SAM103102 Sat Oct 31 01:34:37 1998
 Sample Weight: 13.017 mg
 S98T002955DUP

Curve 1: TGA

File Info: SAM103103 Sat Oct 31 02:35:48 1998

Sample Weight: 14.342 mg

S98T002956

RD MEYERS
PERKIN-ELMER
7 Series Thermal Analysis System
Sat Oct 31 02:49:34 1998

Curve 1: TGA
 File info: SAM103104 Sat Oct 31 03:44:35 1998
 Sample Weight: 15.597 mg
 S98T002956DUP

RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Nov 1 06:10:28 1998

10C/MIN N2
 TEMPERATURE
 0.0 min RATE: 10.0 C/min

LABCORE Data Entry Template for Worklist# 26771

Analyst: DK Instrument: TGA0 3 Book # 11718-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.72</u>	N/A	%
98000509	BY-105 (2)	2 SAMPLE	S98T002957	0	TGA-03	LIQUID	N/A	<u>47.98</u>		%
98000509	BY-105 (2)	3 DUP	S98T002957	0	TGA-03	LIQUID	<u>47.98</u>	<u>49.08</u>	N/A	%
98000509	BY-105 (2)	4 SAMPLE	S98T002958	0	TGA-03	LIQUID	N/A	<u>46.51</u>		%
98000509	BY-105 (2)	5 DUP	S98T002958	0	TGA-03	LIQUID	<u>46.51</u>	<u>45.47</u>	N/A	%
98000509	BY-105 (2)	6 SAMPLE	S98T002959	0	TGA-03	LIQUID	N/A	<u>45.28</u>		%
98000509	BY-105 (2)	7 DUP	S98T002959	0	TGA-03	LIQUID	<u>45.28</u>	<u>44.66</u>	N/A	%

Final page for worklist # 26771

[Signature] 10/24/98
Analyst Signature Date

[Signature] 10/28/98
Analyst Signature Date

Validated 10/29/98 [Signature]

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER102301 Fri Oct 23 20:24:38 1998

Sample Weight: 14.452 mg

117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 534 TO 540.

10C/MIN N2

TEMP: 296.5 °C

TIME: 0.0

min RATE: 10.0 °C/min

Temperature (°C)

J. KING
PERKIN-ELMER7 Series Thermal Analysis System
Fri Oct 23 20:43:25 1998

Curve 1: TGA
 File info: SAM02301 Fri Oct 23 21:57:31 1998
 Sample Weight: 15.326 mg
 S98T002957

100%/MIN N2
 TGA: 35.0 g
 TIME: 555.0 s
 RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 14:08:34 1998

Curve 1: TGA
 File info: SAM102302 Fri Oct 23 23:02:57 1998
 Sample Weight: 16.642 mg
 998T002957 DUP

100/MIN N2
 TGA: 25.0 g
 TEMPE: 500.0 g

0.0 min RATE: 10.0 C/min

Temperature (°C)

RM KINS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 14:18:07 1998

Curve 1: TGA

File Info: SAM102303 Sat Oct 24 00:08:04 1998

Sample Weight: 16.808 mg

S98T002958

10C/MIN N2

TEMP: 250.0 °C

TIME: 0.0

min RATE: 10.0

C/min

PERKIN-ELMER

7 Series Thermal Analysis System

Mon Oct 26 14:19:53 1998

Curve 1: TGA
 File info: SAM102401 Sat Oct 24 01:19:10 1998
 Sample Weight: 15.889 mg
 S98T002958 DUP

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 14:21:41 1998

10C/MIN N2
 THERMO 550.8 C
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: T6A
 File info: SAM102403 Sat Oct 24 02:28:35 1998
 Sample Weight: 17.122 mg
 S98T002959

10C/MIN N2
 TIME: 865.8 s
 RATE: 0.0 min RATE: 10.0 C/min
 RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 14:25:00 1998

Curve 1: TGA

File info: SAM102404 Sat Oct 24 04:22:42 1998

Sample Weight: 16.538 mg

S98T002959 DUP

RM KING
PERKIN-ELMER
7 Series Thermal Analysis System
Mon Oct 26 14:27:00 1998

LABCORE Data Entry Template for Worklist# 26772

Analyst: SLH Instrument: TGA0 3 Book # 117N8A
Method: LA-514-114 Rev/Mod D-1
Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.61</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T002589 0		TGA-03	SOLID	<u>N/I</u>	<u>0</u>		%
98000436	BY-105 (2)	3 DUP	S98T002589 0		TGA-03	SOLID	<u>0</u>	<u>0</u>	<u>N/A</u>	%
98000436	BY-105 (2)	4 SAMPLE	S98T002590 0		TGA-03	SOLID	<u>N/I</u>	<u>2.109</u>		%
98000436	BY-105 (2)	5 DUP	S98T002590 0		TGA-03	SOLID	<u>2.109</u>	<u>2.07</u>	<u>N/A</u>	%

Final page for worklist # 26772

Samuel Hood Boertrunt
Analyst Signature Date 10-29-98

Kim Thomas 10-29-98
Analyst Signature Date

Validated 11/3/98 Michael

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER102802 Wed Oct 28 19: 41: 50 1998
 Sample Weight: 16.294 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 542 TO 544.

Curve 1: TGA
File Info: SAM02808 Wed Oct 28 23:13:12 1998
Sample Weight: 22.088 mg
998T002589

10C/MIN N2
TIME: 500.8 s

SL HOOD BOATRUGHT
PERKIN-ELMER
7 Series Thermal Analysis System
Wed Oct 28 23:19:07 1998

Curve 1: T6A
File Info: SAM02809 Thu Oct 29 00:20:09 1998
Sample Weight: 16.663 mg
S98T002589 DUP

SL HOOD BOATHRIGHT
PERKIN-ELMER
7 Series Thermal Analysis System
Thu Oct 29 00:22:08 1998

10C/MIN N2
TEMP: 200.0 °C
TIME: 9.0 min RATE: 10.0 C/min

Curve 1: TGA
 File info: SAM102901 Thu Oct 29 01:26:57 1998
 Sample Weight: 25.796 mg
 SP8T002590

40C/MIN N2
 TGA: 25.8 g
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: TGA

File Info: SAM102902 Thu Oct 29 02:33:52 1998

Sample Weight: 22.748 mg

S98T002590 DUP

LABCORE Data Entry Template for Worklist# 26773

Analyst: SLH Instrument: TGA0 3 Book # 117N8-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.98</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T002591	0	TGA-03	SOLID	<u>N/A</u>	<u>18.6</u>		%
98000436	BY-105 (2)	3 DUP	S98T002591	0	TGA-03	SOLID	<u>18.6</u>	<u>14.77</u>	<u>N/A</u>	%
98000436	BY-105 (2)	4 SAMPLE	S98T002592	0	TGA-03	SOLID	<u>N/A</u>	<u>2.631</u>		%
98000436	BY-105 (2)	5 DUP	S98T002592	0	TGA-03	SOLID	<u>2.631</u>	<u>2.114</u>	<u>N/A</u>	%

Final page for worklist # 26773

Judith Hord Bischoff
Analyst Signature Date 11-3-98

M. B. Smith
Analyst Signature Date 11/3/98

Validated 11/9/98 [Signature]

Data Entry Comments:
High RPD S98T002592 due to sample inhomogeneities. No rerun. [Signature] 11/9/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER110301 Tue Nov 3 07:01:57 1998
 Sample Weight: 19.677 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 548 TO 553.

SL HOOD-BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 07:03:13 1998

Curve 1: TGA
 File Info: SAM110305 Tue Nov 3 13:36:20 1998
 Sample Weight: 26.341 mg
 S98T002591

Curve 1: TGA
File Info: SAM10306 Tue Nov 3 14: 47: 07 1998
Sample Weight: 28.972 mg
S98T002591 DUP

10C/MIN N2
TEMP: 25.0 C TIME: 0.0 min RATE: 10.0 C/min
SL HOOD BOATRUGHT
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Nov 3 14: 53: 58 1998

Curve 1: T6A
 File Info: SAM10307 Tue Nov 3 15:54:08 1998
 Sample Weight: 27.342 mg
 S98T002592

SL HOOD BOATRUGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Nov 4 13:48:41 1998

Curve 1: TGA

File info: SAM10308 Tue Nov 3 17:40:27 1998

Sample Weight: 29.703 mg

S98T002592 DUP

 10C/MIN N2
 THERM 38.8 g
 TEMPE 503.8 g

0.0 MIN RATE: 10.0 C/MIN

Temperature (°C)

SL HOOD BOATRIGT
PERKIN-ELMER7 Series Thermal Analysis System
Tue Nov 3 15:30:38 1998

LABCORE Data Entry Template for Worklist# 26774

Analyst: SLH Instrument: TGA0 3 Book # 117 N8-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.71</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T002593	0	TGA-03	SOLID	<u>N/A</u>	<u>28.78</u>		%
98000436	BY-105 (2)	3 DUP	S98T002593	0	TGA-03	SOLID	<u>28.78</u>	<u>28.19</u>	<u>N/A</u>	%
98000436	BY-105 (2)	4 SAMPLE	S98T002594	0	TGA-03	SOLID	<u>N/A</u>	<u>32.13</u>		%
98000436	BY-105 (2)	5 DUP	S98T002594	0	TGA-03	SOLID	<u>32.13</u>	<u>33.41</u>	<u>N/A</u>	%

Final page for worklist # 26774

Andrew Hood Boatwright
Analyst Signature Date

10-30-98

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/11/98 JP Hacheler

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER102901 Thu Oct 29 10:35:41 1998

Sample Weight: 14.549 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 554 TO 559

10C/MIN N2
TIME: 30.38
10C/MIN RATE: 10.0 C/min
J1 SOLBRACK
PERKIN-ELMER
7 Series Thermal Analysis System
Thu Oct 29 10:48:35 1998
JTG Solbrack

Curve 1: TGA
 File info: SAM102912 Thu Oct 29 20:41:13 1998
 Sample Weight: 24.738 mg
 S98T002593

SL HODD 804TRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 21:34:17 1998

10C/MIN N2
 TEMPERATURE 350.0 °C
 TIME 8.0 min RATE 10.0 °C/min

Curve 1: TGA
 File Info: SAH102913 Thu Oct 29 22:24:02 1998
 Sample Weight: 24.593 mg
 S98T002593 DUP

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 22:26:07 1998

10C/MIN N2
 THERM: 35.8 g
 TOWEL: 0.0 min RATE: 30.0 C/min

557

Curve 1: TGA
 File info: SAM102814 Thu Oct 28 23:30:05 1998
 Sample Weight: 27.262 mg
 998T002594

10C/MIN N2
 TGA 28.8 g
 TIME: 0.0 MIN RATE: 40.0 C/MIN
 SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 28 23:39 1998

Curve 1: T8A

File info: SAM102915 Fri Oct 30 00:38:46 1998

Sample Weight: 29.440 mg

S98T002594 DUP

10C/MIN N2
 T8A
 29.440 g

TIME: 0.0 min RATE: 10.0 C/min

Temperature (°C)

SI HOD BOATRIGT

PERKIN-ELMER

7 Series Thermal Analysis System

Fri Oct 30 01:00:52 1998

HNF-1663 REV. 0

LABCORE Data Entry Template for Worklist# 26775

Analyst: RDM Instrument: TGA0 3 Book # 11141 Rm 10/31/98

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.80X</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T002711 0		TGA-03	SOLID	<u>N/A</u>	<u>23.70</u>		%
98000436	BY-105 (2)	3 DUP	S98T002711 0		TGA-03	SOLID	<u>23.70</u>	<u>35.90</u>	<u>N/A</u>	%
98000490	BY-105 (2)	4 SAMPLE	S98T002737 0		TGA-03	SOLID	<u>N/A</u>	<u>11.71</u>		%
98000490	BY-105 (2)	5 DUP	S98T002737 0		TGA-03	SOLID	<u>11.71</u>	<u>13.22</u>	<u>N/A</u>	%

Final page for worklist # 26775

RDM 10/31/98
Analyst Signature Date

Kim Thomas 11-10-98
Analyst Signature Date

Validated 11/12/98 Rachelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File Info: TGA103001 Fri Oct 30 08:57:53 1998
 Sample Weight: 19.574 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 560 TO 564.

RM KING *R. King*
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Nov 3 15:30:44 1998

Curve 1: TGA
 File info: SAM103017 Fri Oct 30 22:18:20 1998
 Sample Weight: 8.756 mg
 S98T002711

10C/MIN N2

TEMP: 23.0 °C
 TIME: 0.0 min RATE: 10.0 C/min

RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 22:20:09 1998

Curve 1: TGA
 File info: SAM103018 Fri Oct 30 23:21:03 1998
 Sample Weight: 16.353 mg
 S98T002711DUP

100/MIN N2
 TEMP: 35.0 °C
 TIME: 0.0 min RATE: 10.0 °C/min
 RD MEYERS
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 23:52:34 1998

Curve 1: TGA
 File info: SAM103015 Fri Oct 30 19:36:59 1998
 Sample Weight: 11.055 mg
 S98T002737

10C/MIN N2
 TEMPERATURE: 20.0 °C
 TIME: 0.0 min RATE: 10.0 °C/min
 RD MEYERS
 DUKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 19:57:15 1998

Curve 1: TGA
 File info: SAM103016 Fri Oct 30 20:51:18 1998
 Sample Weight: 14.493 mg
 S98T002737DUP

HNF-1663 REV. 0

100/MIN N2

TEMP: 800.0 °C TIME: 0.0 min RATE: 10.0 °C/min

Temperature (°C)

RO MEYERS
 PERKIN-ELMER

7 Series Thermal Analysis System
 Fri Oct 30 21:22:26 1998

800000000

LABCORE Data Entry Template for Worklist# 26776

Analyst: KRM Instrument: TGA0 3 Bottle # 11708A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.91</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002840	0	TGA-03	SOLID	<u>N/A</u>	<u>26.39</u>		%
98000437	BY-105 (2)	3 DUP	S98T002840	0	TGA-03	SOLID	<u>26.39</u>	<u>32.44</u>	<u>N/A</u>	%
98000437	BY-105 (2)	4. SAMPLE	S98T002841	0	TGA-03	SOLID	<u>N/A</u>	<u>16.08</u>		%
98000437	BY-105 (2)	5 DUP	S98T002841	0	TGA-03	SOLID	<u>16.08</u>	<u>36.75</u>	<u>N/A</u>	%

Final page for worklist # 26776

Analyst Signature [Signature] Date 10-26-98

Analyst Signature [Signature] Date 10-29-98

Validated/ 10/29/98 [Signature]
Rejected

Data Entry Comments:

High RPD S98T002841 due to unknown causes. Rerun requested. [Signature]

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER102601 Mon Oct 26 10:37:06 1998
 Sample Weight: 15.335 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 566 TO 570.

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 10:38:34 1998

10C/MIN N2
 TEMPERATURE 350.0 °C
 TIME 0.0 min RATE 10.0 C/min

Curve 1: TGA
 File Info: SAM102603 Mon Oct 26 14:52:32 1998
 Sample Weight: 16.673 mg
 S98T002840

Curve 1: TGA
 File Info: SAM102604 Mon Oct 26 16:12:59 1998
 Sample Weight: 17.804 mg
 S98T002640 DUP

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 16:30:52 1998

10C/MIN N2
 TIME: 00:08
 RATE: 10.0 C/min

HNF-1663 REV. 0

Curve 1: TGA
File info: SAM102601 Mon Oct 26 11:51:53 1998
Sample Weight: 45.708 mg
S98T002841

KR MONTEITH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 09:53 1998

Curve 1: TGA
 File info: SAM102602 Mon Oct 26 13:42:25 1998
 Sample Weight: 14.410 mg
 S98T002841 DUP

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:11:46 1998

10/30/98 09:18

LABCORE Data Entry Template for Worklist# 26777

Analyst: SLH Instrument: TGA0 3 Book # 117N81AMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.96</u>	<u>N/A</u>	%
98000490	BY-105 (2)	2 SAMPLE	S98T002910	0	TGA-03	SOLID	<u>N/A</u>	<u>5.37</u>	<u>6</u>	<u>K7 10-30-98</u>
98000490	BY-105 (2)	3 DUP	S98T002910	0	TGA-03	SOLID	<u>5.36</u>	<u>10.21</u>	<u>N/A</u>	%
98000490	BY-105 (2)	4 TRIPL	S98T002910	0	TGA-03	SOLID	<u>5.36</u>	<u>9.13</u>	<u>N/A</u>	%
98000490	BY-105 (2)	5 SAMPLE	S98T002911	0	TGA-03	SOLID	<u>N/A</u>	<u>2.99</u>		%
98000490	BY-105 (2)	6 DUP	S98T002911	0	TGA-03	SOLID	<u>2.99</u>	<u>4.01</u>	<u>N/A</u>	%
98000490	BY-105 (2)	7 TRIPL	S98T002911	0	TGA-03	SOLID	<u>2.99</u>	<u>4.32</u>	<u>N/A</u>	%

Final page for worklist # 26777

See Attached for Signature
 Analyst Signature   Date

Kim Thomas 10-30-98
 Analyst Signature   Date

Validated 11/4/98 [Signature]

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26777

Analyst: SLH Instrument: TGA0 3 Bottle # 117108-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.96</u>	N/A	%
98000490	BY-105 (2)	2 SAMPLE	S98T002910 0		TGA-03	SOLID	N/A	<u>5.36</u>		%
98000490	BY-105 (2)	3 DUP 4 TRIPLE	S98T002910 0		TGA-03	SOLID	<u>5.36</u> <u>5.36</u>	<u>10.21</u> <u>9.13</u>	N/A	%
98000490	BY-105 (2)	5A SAMPLE	S98T002911 0		TGA-03	SOLID	N/A	<u>2.99</u>		%
98000490	BY-105 (2)	6A DUP 7 TRIPLE	S98T002911 0		TGA-03	SOLID	<u>2.99</u> <u>2.99</u>	<u>4.01</u> <u>4.32</u>	N/A	%

Final page for worklist # 26777

Sandra H. Hoot
Analyst Signature Date 10-28-98

Analyst Signature Date

Data Entry Comments:

Second STD run due to delay in completion of worklist with a value of 58.61% and a recovery of 98.67%. ~~SLH~~ 11/29/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File Info: TER102702 Tue Oct 27 19:13:30 1998

Sample Weight: 26.908 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 573 TO 580.

573

100%/MIN N2

TEMP: 25.8 °C

TIME: 300.8 s

0.0 min RATE: 10.0 °C/min

SL HOOD BOATRIGT

PERKIN-ELMER

7 Series Thermal Analysis System

Wed Oct 28 01:39:24 1998

Subbed Boattright

Curve 1: TGA
 File Info: TEP102802 Wed Oct 28 19: 41: 50 1998
 Sample Weight: 18.294 mg
 117N8-A

SL HD00 BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 19: 52: 05 1998

James H. Good Boatswain

10C/MIN N2
 TEMP: 26.9 °C
 TIME: 200.0 S
 0.0 min RATE: 10.0 C/min

Curve 1: TBA
 File Info: SAM102802 Wed Oct 28 02:35:23 1998
 Sample Weight: 21.845 mg
 SS8T002910

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 03:02:33 1998

10C/MIN N2
 TBA-2 500.0 g
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: TGA
File info: SAM102803 Wed Oct 28 03:52:55 1998
Sample Weight: 21.192 mg
S98T002910 DUP

SL HOOD BOATRIGHT
PERKIN-ELMER
7 Series Thermal Analysis System
Wed Oct 28 04:11:13 1998

10C/MIN N2
TEMP 500.0 °C
TIME 0.0 min RATE 10.0 C/min

Curve 1: TGA
 File info: SAM02806 Wed Oct 28 20:50:35 1998
 Sample Weight: 25.253 mg
 998T002910 TTP

Curve 1: TGA
 File info: SAM102804 Wed Oct 28 05:02:02 1998
 Sample Weight: 16.117 mg
 998T002911

10C/MIN N2
 THERM 250.0 g
 TIME: 0.0 min RATE: 10.0 C/min
 SL HOOD BOATRIGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 06:53 1998

Curve 1: TGA
 File info: SAM102805 Wed Oct 28 06:07:46 1998
 Sample Weight: 19.502 mg
 S98T002911 DUP

Curve 1: TGA
 File info: SAH102807 Wed Oct 28 21:56:41 1998
 Sample Weight: 16.847 mg
 S98T002911 TPP

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Oct 28 22:20:44 1998

10C/MIN N2
 TIME: 00:08
 RATE: 0.0 MIN RATE: 10.0 C/min

LABCORE Data Entry Template for Worklist# 26778

Analyst: KRM Instrument: TGA0 3 Bottle # 117A8A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.91</u>	<u>N/A</u>	%
98000490	BY-105 (2)	2 SAMPLE	S98T002912	0	TGA-03	SOLID	<u>N/A</u>	<u>1.07</u>		%
98000490	BY-105 (2)	3 DUP	S98T002912	0	TGA-03	SOLID	<u>1.07</u>	<u>1.55</u>	<u>N/A</u>	%
98000490	BY-105 (2)	4 SAMPLE	S98T002913	0	TGA-03	SOLID	<u>N/A</u>	<u>10.28</u>		%
98000490	BY-105 (2)	5 DUP	S98T002913	0	TGA-03	SOLID	<u>10.28</u>	<u>12.31</u>	<u>N/A</u>	%

Final page for worklist # 26778

[Signature] 10-27-98
Analyst Signature Date

Kim Thomas 10-29-98
Analyst Signature Date

[Signature] 10/29/98 Validated

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER102701 Tue Oct 27 06:46:59 1998

Sample Weight: 13.876 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 582 TO 586.

Curve 1: TGA
File info: SAM102706 Tue Oct 27 10:55:32 1998
Sample Weight: 25.474 mg
S98T002912

KR MONTEITH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 12:44 1998

10C/MIN N2
TGA
TIME: 0.0 min RATE: 10.0 C/min
TEMP: 500.0 °C

HNF-1663 REV. 0

Curve 1: TGA
File info: SAM02707 Tue Oct 27 13:02:39 1988
Sample Weight: 10.261 mg
S98T002912 DUP

KR MONTEITH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 13:04:21 1988

10C/MIN N2
TIME: 35.0 S
TEMP: 500.0 C
0.0 min RATE: 10.0 C/min

Curve 1: TGA
File info: SAM102708 Tue Oct 27 14:09:47 1998
Sample Weight: 20.404 mg
S98T002913

KR MONTEITH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 14:10:43 1998

Curve 1: TGA

File info: SAM102709 Tue Oct 27 15:17:48 1998

Sample Weight: 30.564 mg

S98T002913 DUP

LABCORE Data Entry Template for Worklist# 26779

Analyst: KRM Instrument: TGA0 3 Book # 117 N8AMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.91</u>	<u>N/A</u>	%
98000490	BY-105 (2)	2 SAMPLE	S98T002914	0	TGA-03	SOLID	<u>N/A</u>	<u>15.86</u>		%
98000490	BY-105 (2)	3 DUP	S98T002914	0	TGA-03	SOLID	<u>15.86</u>	<u>14.90</u>	<u>N/A</u>	%
98000509	BY-105 (2)	4 SAMPLE	S98T002980	0	TGA-03	SOLID	<u>N/A</u>	<u>30.89</u>		%
98000509	BY-105 (2)	5 DUP	S98T002980	0	TGA-03	SOLID	<u>30.89</u>	<u>23.70</u>	<u>N/A</u>	%

Final page for worklist # 26779

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File Info: TER102701 Tue Oct 27 06:46:59 1998

Sample Weight: 13.876 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 588 TO 589.

10C/MIN N2

TEMP: 300.0 °C

TIME: 0.0

min RATE: 10.0 °C/min

Temperature (°C)

KR MONTEITH
PERKIN-ELMER

7 Series Thermal Analysis System
Tue Oct 27 06:52:38 1998

Curve 1: TGA
 File Info: SAK102711 Tue Oct 27 16:27:55 1998
 Sample Weight: 28.048 mg
 S98T002914

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 16:28:52 1998

10C/MIN N2
 0.0 min RATE: 10.0 C/min

TEMP: 80.0 °C
 TIME: 0.0

Curve 1: TGA
File Info: SAM102712 Tue Oct 27 17:35:10 1998
Sample Weight: 21.297 mg
S98T002914 DUP

KR MONTEITH
PERKIN-ELMER
7 Series Thermal Analysis System
Tue Oct 27 18:41:06 1998

100/MIN N2
35.0 g
TIME: 500.0 g
0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File info: SAM102704 Tue Oct 27 07:54:11 1998
 Sample Weight: 10.347 mg
 S98T002980

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 08:58:58 1998

10C/MIN N2
 TEMPS 35.0 C
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File Info: SAM102705 Tue Oct 27 09:47:21 1998
 Sample Weight: 27.514 mg
 S98T002980 DUP

10C/MIN N2
 THERM 500.0 g
 TIME: 0.0 min RATE: 10.0 C/min

KR MONTEITH
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:59:11 1998

worklistrpt Version 2.1 05/15/95
10/23/98 10:05

LABCORE Data Entry Template for Worklist# 26780

Analyst: NA Instrument: TGA0 3 Boole # 11718-AMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>59.91</u>	<u>N/A</u>	%
98000509	BY-105 (2)	2 SAMPLE	S98T002981 0		TGA-03	SOLID	<u>N/A</u>	<u>34.57</u>	<u>11.77</u>	%
98000509	BY-105 (2)	3 DUP	S98T002981 0		TGA-03	SOLID	<u>34.57</u>	<u>45.20</u>	<u>N/A</u>	%
98000509	BY-105 (2)	4 SAMPLE	S98T002982 0		TGA-03	SOLID	<u>N/A</u>	<u>11.77</u>	<u>N/A</u>	%
98000509	BY-105 (2)	5 DUP	S98T002982 0		TGA-03	SOLID	<u>11.77</u>	<u>10.07</u>	<u>N/A</u>	%
98000509	BY-105 (2)	6 SAMPLE	S98T002983 0		TGA-03	SOLID	<u>N/A</u>	<u>36.45</u>	<u>N/A</u>	%
98000509	BY-105 (2)	7 DUP	S98T002983 0		TGA-03	SOLID	<u>36.45</u>	<u>43.77</u>	<u>N/A</u>	%
8 TRIPL							<u>36.45</u>	<u>35.66</u>		

Final page for worklist # 26780

Not Kim 10/24/98
Analyst Signature Date

Kim Thomas 10-29-98
Analyst Signature Date

Validated to 10/24/98 Dr. Scheller

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TERA102601 Mon Oct 26 10:37:06 1998

Sample Weight: 15.335 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 594 TO 601.

10C/MIN N2

TEMP: 33.8 °C

TIMER1: 0.0 min RATE1: 10.0 C/min

Temperature (°C)

10/27/19
~~KP-MONTITH~~

PERKIN-ELMER

Series Thermal Analysis system
Tue Oct 27 02:05:45 1998

Curve 1: TGA
 File Info: SAM102605 Mon Oct 26 19:43:14 1998
 Sample Weight: 28.255 mg
 S98T002981

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:13:44 1998

10C/MIN N2
 TEMP: 35.8 °C
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File info: SAM102606 Mon Oct 26 20:49:14 1998
 Sample Weight: 14.568 mg
 S98T002981 DUP

RM KING
 PEKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:21:31 1998

Curve 1: TGA
 File Info: SAM102607 Mon Oct 26 22:05:43 1998
 Sample Weight: 16.218 mg
 598T002982

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Mon Oct 26 22:09:13 1998

10C/MIN N2
 TEMPERATURE 35.0 °C
 TIME 500.0 S
 0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File info: SAM102608 Mon Oct 26 23:11:21 1998
 Sample Weight: 8.625 mg
 S98T002982 DUP

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 08:25:44 1998

100/MIN N2
 THERM 555.8 S
 TIME 0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File info: SAM102609 Tue Oct 27 00:47:07 1998
 Sample Weight: 13.780 mg
 S98T002983

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 09:27:45 1998

Curve 1: TSA
 File info: SAM102701 Tue Oct 27 02:51:10 1998
 Sample Weight: 13.290 mg
 S987002983 DUP

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Tue Oct 27 08:28:35 1998

10C/MIN N2
 TEMP: 85.8 °C
 TIME: 0.0 min RATE: 10.0 C/min

Curve 1: TGA

File Info: SAM02703 Tue Oct 27 04:25:02 1998

Sample Weight: 15.655 mg

S98T002983 TRIP

LABCORE Data Entry Template for Worklist# 26851

Analyst: DK Instrument: TGA0 3 Book # 117178-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.80</u>	<u>N/A</u>	%
98000510	BY-105 (2)	2 SAMPLE	S98T003002	0	TGA-03	LIQUID	<u>N/A</u>	<u>53.24</u>		%
98000510	BY-105 (2)	3 DUP	S98T003002	0	TGA-03	LIQUID	<u>53.24</u>	<u>51.69</u>	<u>N/A</u>	%
98000510	BY-105 (2)	4 SAMPLE	S98T003003	0	TGA-03	LIQUID	<u>N/A</u>	<u>51.64</u>		%
98000510	BY-105 (2)	5 DUP	S98T003003	0	TGA-03	LIQUID	<u>51.64</u>	<u>51.42</u>	<u>N/A</u>	%

Final page for worklist # 26851

Mark M. [Signature] 10/30/98
Analyst Signature Date

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/11/98 [Signature]

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER105001 Fri Oct 30 08:57:33 1998

Sample Weight: 19.574 mg

117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 603 TO 607

RW KING
 DEPT. 10-10-98

7 Series Thermal Analysis System
 Fri Oct 30 13:20:16 1998

Curve 1: TGA

File Info: SAM103009 Fri Oct 30 14:51:07 1998

Sample Weight: 15.001 mg

S98T003002

10C/MIN N2

TEMP: 30.0 °C TIME: 0.0 min RATE: 10.0 C/min

Temperature (°C)

 RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 14:55:11 1998

Curve 1: TGA
 File Info: SAM103010 Fri Oct 30 15:56:46 1998
 Sample Weight: 14.872 mg
 S98T003002 DUP

10C/MIN N2

TEMP: 35.0 °C
 TIME: 0.0 min RATE: 10.0 C/min

Temperature (°C)

RW KING
 PEAKIN-ELMER

7 Series Thermal Analysis System
 Fri Oct 30 15:57:03 1998

Curve 1: TGA
 File Info: SAM103011 Fri Oct 30 17:02:22 1998
 Sample Weight: 16.454 mg
 S98T003003

RM KING
 DRYAN-ELMER
 7 Series Thermal Analysis System
 Fri Oct 30 17:04:04 1998

Curve 1: TGA
 File Info: SAM103012 Fri Oct 30 18: 08: 04 1998
 Sample Weight: 16.364 mg
 S98T003003 DUP

worklistrpt Version 2.1 05/15/95
10/26/98 14:35

Page: 1

LABCORE Data Entry Template for Worklist# 26853

Analyst: SLH Instrument: TGA0 3 Boil # 117N8-AMethod: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.61</u>	<u>N/A</u>	%
98000510	BY-105 (2)	2 SAMPLE	S98T003004	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.58</u>		%
98000510	BY-105 (2)	3 DUP	S98T003004	0	TGA-03	LIQUID	<u>52.58</u>	<u>52.42</u>	<u>N/A</u>	%
98000510	BY-105 (2)	4 SAMPLE	S98T003005	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.14</u>		%
98000510	BY-105 (2)	5 DUP	S98T003005	0	TGA-03	LIQUID	<u>52.14</u>	<u>52.09</u>	<u>N/A</u>	%

Final page for worklist # 26853

Sandra H. Hines Boatright
Analyst Signature Date

10-29-98

Kim Thomas 11-5-98
Analyst Signature Date

Validated 11/11/98 P. Schelor

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER102802 Wed Oct 28 19: 41: 50 1998
 Sample Weight: 18.294 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1007 TO 1013.

Curve 1: TGA

File Info: SAM102903 Thu Oct 29 03:40:54 1998

Sample Weight: 15.560 mg

S98T003004

610

10C/MIN N2

TIME: 35.9 s

10C/MIN N2

0.0 min RATE: 10.0 C/min

Temperature (°C)

SL HOOD BOATRIGT
PERKIN-ELMER7 Series Thermal Analysis System
Thu Oct 29 03:43:16 1998

Curve 1: TGA
 File info: SAM102904 Thu Oct 29 04:47:24 1998
 Sample Weight: 15.419 mg
 S98T003004 DUP

SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 04:57:14 1998

10C/MIN N2
 THERM: 35.8 g
 TGA: 35.8 g
 0.0 min RATE: 10.0 C/min

Curve 1: TGA
 File Info: SAM102905 Thu Oct 29 05:53:12 1998
 Sample Weight: 17.409 mg
 S98T003005

612

Curve 1: TGA
 File Info: SAM102908 Thu Oct 29 12:04:36 1998
 Sample Weight: 13.861 mg
 S98T003005DUP

10C/MIN N2
 TIME: 25.8 8
 RATE: 0.0 min RATE: 10.0 C/min
 JI SOL BRACK
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Thu Oct 29 12:45:25 1998

LABCORE Data Entry Template for Worklist# 26912

Analyst: RLK Instrument: TGA0 3 Book # 11708-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: Rerun due to high RPD BY-105, KNT

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>58.98</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T002841 0		TGA-03	SOLID	<u>N/A</u>	<u>26.71</u>		%
98000437	BY-105 (2)	3 DUP	S98T002841 0		TGA-03	SOLID	<u>26.71</u>	<u>15.01</u>	<u>N/A</u>	%

Final page for worklist # 26912

RLK 11/14/98
Analyst Signature Date

MCB 11/19/98
Analyst Signature Date

Validated 11/19/98 Brachet

Data Entry Comments:

Reproducible high RPD. No further rerun. 11/19/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER110301 Tue Nov 3 07:01:57 1998
 Sample Weight: 19.677 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 615 TO 617.

Curve 1: TGA
 File info: SAM110309 Tue Nov 3 20:23:13 1998
 Sample Weight: 15.550 mg
 S98T002841

HNF-1663 REV. 0

Curve 1: TGA
File info: SAM10310 Tue Nov 3 21:49:21 1998
Sample Weight: 19.851 mg
S98T002841 DUP

LABCORE Completed Worklist Report for Worklist# 27637

Analyst: slh

Instrument: TGA03

Book#: 117N8A

Method: LA-514-114 Rev/Mod _____

Worklist Comment: BY-105 TGA-03 RLC

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0		TGA-03	LIQUID	5.94e1	58.80	98.990 % Recovery	
2	SAMPLE	S98T003383	0		TGA-03	LIQUID	N/A	52.12	%	
3	DUP	S98T003383	0		TGA-03	LIQUID	52.12	52.68	1.069 RPD	

Final page for worklist# 27637

Analyst Signature _____ Date _____

Mary Tracy 12-31-98
Analyst Signature Date

Rachel 1/4/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27637

Analyst: SLH Instrument: TGA0 3 Book # 117N8-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 TGA-03 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.80</u>	<u>N/A</u>	%
98000437	BY-105 (2)	2 SAMPLE	S98T003383	0	TGA-03	LIQUID	<u>N/A</u>	<u>52.12</u>		%
98000437	BY-105 (2)	3 DUP	S98T003383	0	TGA-03	LIQUID	<u>52.12</u>	<u>52.68</u>	<u>N/A</u>	%

Final page for worklist # 27637

Jandra Hard Boekwert
Analyst Signature Date
12-20-98

Mary Tracy 12-31-98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER122001 Sun Dec 20 07: 48: 36 1998
 Sample Weight: 20.284 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 620 TO 622

10C/MIN N2
 TEMP: 35.0 °C
 TIME: 300.0 S
 10C/MIN RATE: 10.0 °C/min
 SL HOOD BOATRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 08: 52 1998

Randall Wood Boatright

Curve 1: TGA
 File Info: SAM122004 Sun Dec 20 09: 01: 10 1998
 Sample Weight: 13.038 mg
 S98T003363

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 09: 01: 10 1998

100/MIN N2
 TEMP: 535.8 °C
 TIME: 0.0 MIN RATE: 10.0 °C/MIN

Curve 1: TGA
 File info: SAM122005 Sun Dec 20 10:48:33 1998
 Sample Weight: 15.709 mg
 S98T003383 DUP

SL HOOD BOATRRIGHT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sun Dec 20 10:58:40 1998

100./MIN N2
 35.8 g
 0.0 min RATE: 10.0 C/min
 TIME: 55.8 g

LABCORE Completed Worklist Report for Worklist# 27638

Analyst: rwk

Instrument: TGA03

Book#: 117N8A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 TGA-03 RLC

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD		0	TGA-03	SOLID	5.94e1	60.65	*	102.121 % Recovery
2	SAMPLE	S98T003391	0	TGA-03	SOLID	N/A	2.53		%
3	DUP	S98T003391	0	TGA-03	SOLID	2.53	2.21		13.502 RPD
4	SAMPLE	S98T003396	0	TGA-03	SOLID	N/A	31.28		%
5	DUP	S98T003396	0	TGA-03	SOLID	31.28	29.96		4.311 RPD

Final page for worklist# 27638

Analyst Signature

Date

Mary J. Swann
Analyst Signature

12-29-98
Date

D. Schelcher
Reviewer Signature

12/30/98
Date

LABCORE Data Entry Template for Worklist# 27638

Analyst: PK Instrument: TGA0 3 Book # 11708-A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 TGA-03 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>60.66</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T003391	0	TGA-03	SOLID	<u>N/A</u>	<u>2.53</u>		%
98000436	BY-105 (2)	3 DUP	S98T003391	0	TGA-03	SOLID	<u>2.53</u>	<u>2.21</u>	<u>N/A</u>	%
98000436	BY-105 (2)	4 SAMPLE	S98T003396	0	TGA-03	SOLID	<u>N/A</u>	<u>31.28</u>		%
98000436	BY-105 (2)	5 DUP	S98T003396	0	TGA-03	SOLID	<u>31.28</u>	<u>29.96</u>	<u>N/A</u>	%

Final page for worklist # 27638

Mohd Nizam 12/20/98
Analyst Signature Date

Mary Jane 12-29-98
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA

File info: TER121998 Sat Dec 19 19:04:10 1998

Sample Weight: 19.974 mg

117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 625 TO 629

10C/MIN N2

TIME: 300.0 s

min RATE: 10.0 C/min

Temperature (°C)

RW KING
PERKIN-ELMER
7 Series Thermal Analysis System
Sat Dec 19 22:50:49 1998

Curve 1: T6A
 File Info: SAM121903 Sat Dec 19 22:22:22 1998
 Sample Weight: 15.807 mg
 S98T003391

Curve 1: TGA
 File Info: SAM122001 Sun Dec 20 01:06:09 1998
 Sample Weight: 13.974 mg
 S98T003391 DUP

Curve 1: TGA
 File info: SAM122002 Sun Dec 20 02:14:29 1998
 Sample Weight: 24.371 mg
 S98T003396

Curve 1: TSA
 File info: SAM12003 Sun Dec 20 03:19:35 1998
 Sample Weight: 17.773 mg
 S98T003396 DUP

LABCORE Completed Worklist Report for Worklist# 27644

Analyst: rwk

Instrument: TGA03

Book#: 117N8A

Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		TGA-03	SOLID	5.94e1	60.66	102.121 % Recovery	
2 SAMPLE	S98T003386	0	TGA-03	SOLID	N/A	3.42	%	
3 DUP	S98T003386	0	TGA-03	SOLID	3.42	3.50	2.312 RPD	

Final page for worklist# 27644

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 27644

Analyst: DK Instrument: TGA0 3 Book # 11778-4Method: LA-514-114 Rev/Mod D-1

Worklist Comment: BY-105 FOR TGA-03 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	SOLID	<u>59.4</u>	<u>60.66</u>	<u>N/A</u>	%
98000490	BY-105 (2)	2 SAMPLE	S98T003386	0	TGA-03	SOLID	<u>N/A</u>	<u>3.42</u>		%
98000490	BY-105 (2)	3 DUP	S98T003386	0	TGA-03	SOLID	<u>3.42</u>	<u>3.50</u>	<u>N/A</u>	%

Final page for worklist # 27644

Analyst Signature Mob. King Date 12/14/98Analyst Signature Mary F. King Date 12-29-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File Info: TER121998 Sat Dec 19 19:04:10 1998
 Sample Weight: 19.974 mg
 117N8-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 ANALYZED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 632 TO 634.

632

10C/MIN N2
 TGA 300.0 g
 TIME: 0.0 min RATE: 10.0 C/min
 RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sat Dec 19 19:04:44 1998

Curve 1: TGA
 File Info: SAM121901 Sat Dec 19 20:07:08 1998
 Sample Weight: 13.488 mg
 S98T003386

10C/MIN N2
 TEMP: 35.0 °C
 TIME: 566.0 s
 0.0 min RATE: 10.0 C/min
 Temperature (°C)
 RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sat Dec 19 20:14:45 1998

Curve 1: TGA
 File Info: S4121902 Sat Dec 19 21:13:15 1998
 Sample Weight: 11.656 mg
 S98T003386 DUP

RM KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Sat Dec 19 21:26:24 1998

10C/MIN N2
 TEMPE 25.0 C
 TIME: 0.0 min RATE: 10.0 C/min

LABCORE Completed Worklist Report for Worklist# 28072

Analyst: rwk

Instrument: TGA03

Book#: 117N8A

Method: LA-514-114 Rev/Mod D-2

Worklist Comment: BY105 FB FOR TGA-03 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		TGA-03	LIQUID	5.94e1	58.62	98.687 % Recovery	
2 SAMPLE	S98T002391	0	TGA-03	LIQUID	N/A	99.57	%	
3 DUP	S98T002391	0	TGA-03	LIQUID	99.57	99.45	0.121 RPD	

Final page for worklist# 28072

Analyst Signature

Date

[Signature]
Analyst Signature

1/25/99
Date

[Signature]
Reviewer Signature

2/1/99
Date

LABCORE Data Entry Template for Worklist# 28072 Analyst: PK Instrument: TGA0 03 Book # 117787AMethod: LA-514-114 Rev/Mod D-2

Worklist Comment: BY105 FB FOR TGA-03 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			TGA-03	LIQUID	<u>59.4</u>	<u>58.62</u>	<u>N/A</u>	%
98000436	BY-105 (2)	2 SAMPLE	S98T002391	0	TGA-03	LIQUID	<u>N/A</u>	<u>99.57</u>		%
98000436	BY-105 (2)	3 DUP	S98T002391	0	TGA-03	LIQUID	<u>99.57</u>	<u>99.45</u>	<u>N/A</u>	%

Final page for worklist # 28072 Analyst Signature PK Date 1/27/99Analyst Signature C.L. O'Brien Date 1/29/99

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

Curve 1: TGA
 File info: TER012701 Wed Jan 27 07:04:58 1999
 Sample Weight: 19.417 mg
 117NB-A

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 637 TO 639.

10C/MIN N2
 TGA 38.8 g
 TGA 388.8 g
 TIME: 0.0 min RATE: 10.0 g/min
 SL HOOD BOATRGT
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Jan 27 22:28:00 1999

Curve 1: TGA
File info: SAN012711 Wed Jan 27 21:56:24 1999
Sample Weight: 9.666 mg
9997002391

Curve 1: TGA
 File Info: SAN012712 Wed Jan 27 23:08:08 1999
 Sample Weight: 10.251 mg
 S99T002391 DUP

RW KING
 PERKIN-ELMER
 7 Series Thermal Analysis System
 Wed Jan 27 23:36:04 1999

100./MIN N2
 TEMP: 555.0 °C
 TIME: 0.0 min RATE: 10.0 °C/min

LABCORE Completed Worklist Report for Worklist# 28183

Analyst: **rwk**

Instrument: **BA001**

Book#: **134N16C**

Method: **LA-510-112 Rev/Mod** _____

Worklist Comment: **BY105 FOR SPG-01 RLC**

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		SPG-01	LIQUID	1.3779	1.364	98.991 % Recovery	
2 SAMPLE	S98T002806	0	SPG-01	LIQUID	N/A	1.454	1.00e-003	Sp.G.
3 DUP	S98T002806	0	SPG-01	LIQUID	1.454	1.436	1.246 RPD	
4 SAMPLE	S98T002807	0	SPG-01	LIQUID	N/A	1.424	1.00e-003	Sp.G.
5 DUP	S98T002807	0	SPG-01	LIQUID	1.424	1.440	1.117 RPD	
6 SAMPLE	S98T002808	0	SPG-01	LIQUID	N/A	1.444	1.00e-003	Sp.G.
7 DUP	S98T002808	0	SPG-01	LIQUID	1.444	1.439	0.347 RPD	
8 SAMPLE	S98T002809	0	SPG-01	LIQUID	N/A	1.445	1.00e-003	Sp.G.
9 DUP	S98T002809	0	SPG-01	LIQUID	1.445	1.433	0.834 RPD	

Final page for worklist# 28183

Analyst Signature _____

Date _____

Mary Truong 2-9-99
Analyst Signature Date

B. Rachel 2/9/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28183

Analyst: PK Instrument: BA001 Book # 134716-C

Method: LA-510-112 Rev/Mod E0

Worklist Comment: BY105 FOR SPG-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	2 SAMPLE	S98T002806	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	3 DUP	S98T002806	0	SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	4 SAMPLE	S98T002807	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	5 DUP	S98T002807	0	SPG-01	LIQUID			N/A	Sp.G.
98C00437	BY-105 (2)	6 SAMPLE	S98T002808	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	7 DUP	S98T002808	0	SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	8 SAMPLE	S98T002809	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	9 DUP	S98T002809	0	SPG-01	LIQUID			N/A	Sp.G.

Final page for worklist # 28183

Mike King 2/4/99
Analyst Signature Date

Mary Franz 2-9-99
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		STD	REPLICATE
STD	Gross Weight (W2)	1.9918	2.0396
Work List	Tare Weight (W1)	1.8559	1.9026
28183	Weight of Solution (W2-W1)	0.1359	0.137
Test Code	Volume of Solution μ L	100.0000	100.0000
SPG-01	Specific Gravity	1.3590	1.3700
Matrix	Specific Gravity (Average)	1.3645	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity Average =	1.364	
03:14 PM			

Data Entry by:

Date:

02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0486	
Work List	Tare Weight (W1)	1.9032	
28183	Weight of Solution (W2-W1)	0.1454	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4540	NA
Matrix			
LIQUID			
Sample #			
S98T002806			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.454	
03:14 PM			

Data Entry by:	Date:	02/09/99
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0049	
Work List	Tare Weight (W1)	1.8613	
28183	Weight of Solution (W2-W1)	0.1436	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4360	NA
Matrix			
LIQUID			
Sample #			
S98T002806			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.436	
03:14 PM			

Data Entry by:

Date: 02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0226	
Work List	Tare Weight (W1)	1.8802	
28183	Weight of Solution (W2-W1)	0.1424	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4240	NA
Matrix			
LIQUID			
Sample #			
S98T002807			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.424	
03:14 PM			

Data Entry by: _____ Date: 02/09/99

Approved by: *NA* _____ Date: _____

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0199	
Work List	Tare Weight (W1)	1.8759	
28183	Weight of Solution (W2-W1)	0.144	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4400	NA
Matrix			
LIQUID			
Sample #			
S98T002807			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.440	
03:14 PM			

Data Entry by:	<i>NA</i>	Date:	02/09/99
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9367	
Work List	Tare Weight (W1)	1.7923	
28183	Weight of Solution (W2-W1)	0.1444	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4440	NA
Matrix			
LIQUID			
Sample #			
S98T002808			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.444	
03:14 PM			

Data Entry by:	Date: 02/09/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0062	
Work List	Tare Weight (W1)	1.8623	
28183	Weight of Solution (W2-W1)	0.1439	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4390	NA
Matrix			
LIQUID			
Sample #			
S98T002808			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.439	
03:14 PM			

Data Entry by:	Date:	02/09/99
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0323	
Work List	Tare Weight (W1)	1.8878	
28183	Weight of Solution (W2-W1)	0.1445	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4450	NA
Matrix			
LIQUID			
Sample #			
S98T002809			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.445	
03:14 PM			

Data Entry by:

Date:

02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.9924	
Work List	Tare Weight (W1)	1.8491	
28183	Weight of Solution (W2-W1)	0.1433	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4330	NA
Matrix			
LIQUID			
Sample #			
S98T002809			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.433	
03:14 PM			

Data Entry by:

Date:

02/09/99

Approved by:

Date:

LABCORE Completed Worklist Report for Worklist# 28184

Analyst: krm

Instrument: BA001

Book#: 134N16C

Method: LA-510-112 Rev/Mod _____

Worklist Comment: BY105 FOR SPG-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		SPG-01	LIQUID	1.3779	1.365	99.064 % Recovery	
2 SAMPLE	S98T002810	0	SPG-01	LIQUID	N/A	1.399	1.00e-003	Sp.G.
3 DUP	S98T002810	0	SPG-01	LIQUID	1.399	1.411	0.854 RPD	
4 SAMPLE	S98T002811	0	SPG-01	LIQUID	N/A	1.429	1.00e-003	Sp.G.
5 DUP	S98T002811	0	SPG-01	LIQUID	1.429	1.428	0.070 RPD	
6 SAMPLE	S98T002812	0	SPG-01	LIQUID	N/A	1.434	1.00e-003	Sp.G.
7 DUP	S98T002812	0	SPG-01	LIQUID	1.434	1.417	1.193 RPD	
8 SAMPLE	S98T002813	0	SPG-01	LIQUID	N/A	1.453	1.00e-003	Sp.G.
9 DUP	S98T002813	0	SPG-01	LIQUID	1.453	1.376	5.444 RPD	

Final page for worklist# 28184

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 28184

Analyst: ARM Instrument: BA001 Book # 134216C

Method: LA-510-112 Rev/Mod E-0

Worklist Comment: BY105 FOR SPG-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	2 SAMPLE	S98T002810	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	3 DUP	S98T002810	0	SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	4 SAMPLE	S98T002811	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	5 DUP	S98T002811	0	SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	6 SAMPLE	S98T002812	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	7 DUP	S98T002812	0	SPG-01	LIQUID			N/A	Sp.G.
98000437	BY-105 (2)	8 SAMPLE	S98T002813	0	SPG-01	LIQUID	N/A			Sp.G.
98000437	BY-105 (2)	9 DUP	S98T002813	0	SPG-01	LIQUID			N/A	Sp.G.

Final page for worklist # 28184

Hut Mita 2-6-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		STD	REPLICATE
STD	Gross Weight (W2)	1.4956	1.4441
Work List	Tare Weight (W1)	1.3589	1.3077
28184	Weight of Solution (W2-W1)	0.1367	0.1364
Test Code	Volume of Solution μ L	100.0000	100.0000
SPG-01	Specific Gravity	1.3670	1.3640
Matrix	Specific Gravity (Average)	1.3655	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity Average =	1.365	
12:00 AM			

Data Entry by:

Date: 02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4980	
Work List	Tare Weight (W1)	1.3581	
28184	Weight of Solution (W2-W1)	0.1399	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.3990	NA
Matrix			
LIQUID			
Sample #			
S98T002810			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/06/99	v RESULT v		
Time	Specific Gravity =	1.399	
12:00 AM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.4673	
Work List	Tare Weight (W1)	1.3262	
28184	Weight of Solution (W2-W1)	0.1411	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4110	NA
Matrix			
LIQUID			
Sample #			
S98T002810			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.411	
12:00 AM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4825	
Work List	Tare Weight (W1)	1.3396	
28184	Weight of Solution (W2-W1)	0.1429	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4290	NA
Matrix			
LIQUID			
Sample #			
S98T002811			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.429	
12:00 AM			

Data Entry by:	Date:	02/10/99
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.4952	
Work List	Tare Weight (W1)	1.3524	
28184	Weight of Solution (W2-W1)	0.1428	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4280	NA
Matrix			
LIQUID			
Sample #			
S98T002811			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.428	
12:00 AM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4320	
Work List	Tare Weight (W1)	1.2886	
28184	Weight of Solution (W2-W1)	0.1434	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4340	NA
Matrix			
LIQUID			
Sample #			
S98T002812			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.434	
12:00 AM			

Data Entry by:

Date:

02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.4687	
Work List	Tare Weight (W1)	1.3270	
28184	Weight of Solution (W2-W1)	0.1417	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4170	NA
Matrix			
LIQUID			
Sample #			
S98T002812			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.417	
12:00 AM			

Data Entry by:

Date:

02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.5054	
Work List	Tare Weight (W1)	1.3601	
28184	Weight of Solution (W2-W1)	0.1453	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4530	NA
Matrix			
LIQUID			
Sample #			
S98T002813			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.453	
12:00 AM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.4739	
Work List	Tare Weight (W1)	1.3363	
28184	Weight of Solution (W2-W1)	0.1376	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.3760	NA
Matrix			
LIQUID			
Sample #			
S98T002813			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/06/99	v RESULT v		
Time	Specific Gravity =	1.376	
12:00 AM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

LABCORE Completed Worklist Report for Worklist# 28185

Analyst: rwk

Instrument: BA001

Book#: 134N16C

Method: LA-510-112 Rev/Mod _____

Worklist Comment: BY105 FOR SPG-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	SPG-01	LIQUID	1.3779	1.363	98.919 % Recovery	
2 SAMPLE	S98T002938 0	SPG-01	LIQUID	N/A	1.467	1.00e-003	Sp.G.
3 DUP	S98T002938 0	SPG-01	LIQUID	1.467	1.457	0.684 RPD	
4 SAMPLE	S98T002955 0	SPG-01	LIQUID	N/A	1.426	1.00e-003	Sp.G.
5 DUP	S98T002955 0	SPG-01	LIQUID	1.426	1.432	0.420 RPD	
6 SAMPLE	S98T002956 0	SPG-01	LIQUID	N/A	1.440	1.00e-003	Sp.G.
7 DUP	S98T002956 0	SPG-01	LIQUID	1.440	1.433	0.487 RPD	
8 SAMPLE	S98T002957 0	SPG-01	LIQUID	N/A	1.484	1.00e-003	Sp.G.
9 DUP	S98T002957 0	SPG-01	LIQUID	1.484	1.438	3.149 RPD	

Final page for worklist# 28185

Analyst Signature _____ Date _____

Mary Jones 2-9-99
Analyst Signature _____ Date _____

Rachelor 2/9/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28185 Analyst: MLK Instrument: BA001 Book # 134716-CMethod: LA-510-112 Rev/Mod E-D

Worklist Comment: BY105 FOR SPG-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID			N/A	Sp.G.
98000490	BY-105 (2)	2 SAMPLE	S98T002938	0	SPG-01	LIQUID	N/A			Sp.G.
98000490	BY-105 (2)	3 DUP	S98T002938	0	SPG-01	LIQUID			N/A	Sp.G.
98000509	BY-105 (2)	4 SAMPLE	S98T002955	0	SPG-01	LIQUID	N/A			Sp.G.
98000509	BY-105 (2)	5 DUP	S98T002955	0	SPG-01	LIQUID			N/A	Sp.G.
98000509	BY-105 (2)	6 SAMPLE	S98T002956	0	SPG-01	LIQUID	N/A			Sp.G.
98000509	BY-105 (2)	7 DUP	S98T002956	0	SPG-01	LIQUID			N/A	Sp.G.
98000509	BY-105 (2)	8 SAMPLE	S98T002957	0	SPG-01	LIQUID	N/A			Sp.G.
98000509	BY-105 (2)	9 DUP	S98T002957	0	SPG-01	LIQUID			N/A	Sp.G.

Final page for worklist # 28185

M. L. King 2/4/99
 Analyst Signature Date

Mary Franz 2-9-99
 Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		STD	REPLICATE
STD	Gross Weight (W2)	2.0385	1.9170
Work List	Tare Weight (W1)	1.9029	1.7800
28185	Weight of Solution (W2-W1)	0.1356	0.137
Test Code	Volume of Solution μ L	100.0000	100.0000
SPG-01	Specific Gravity	1.3560	1.3700
Matrix	Specific Gravity (Average)	1.3630	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity Average =	1.363	
03:14 PM			

Data Entry by:	Date:	02/09/99
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0321	
Work List	Tare Weight (W1)	1.8854	
28185	Weight of Solution (W2-W1)	0.1467	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4670	NA
Matrix			
LIQUID			
Sample #			
S98T002938			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.467	
03:14 PM			

Data Entry by:	Date:	02/09/99
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	2.0313	
Work List	Tare Weight (W1)	1.8856	
28185	Weight of Solution (W2-W1)	0.1457	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4570	NA
Matrix			
LIQUID			
Sample #			
S98T002938			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.457	
03:14 PM			

Data Entry by: _____ Date: 02/09/99

Approved by: NA Date: _____

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9940	
Work List	Tare Weight (W1)	1.8514	
28185	Weight of Solution (W2-W1)	0.1426	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4260	NA
Matrix			
LIQUID			
Sample #			
S98T002955			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \times 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \times 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.426	
03:14 PM			

Data Entry by:

Date: 02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.9974	
Work List	Tare Weight (W1)	1.8542	
28185	Weight of Solution (W2-W1)	0.1432	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4320	NA
Matrix			
LIQUID			
Sample #			
S98T002955			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.432	
03:14 PM			

Data Entry by: _____ Date: 02/09/99

Approved by: *NA* _____ Date: _____

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9640	
Work List	Tare Weight (W1)	1.8200	
28185	Weight of Solution (W2-W1)	0.144	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4400	NA
Matrix			
LIQUID			
Sample #			
S98T002956			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99			
Time			
03:14 PM			

v RESULT v	
Specific Gravity =	1.440

Data Entry by:

Date: 02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.9994	
Work List	Tare Weight (W1)	1.8561	
28185	Weight of Solution (W2-W1)	0.1433	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4330	NA
Matrix			
LIQUID			
Sample #			
S98T002956			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	$\vee$ RESULT $\vee$		
Time	Specific Gravity =	1.433	
03:14 PM			

Data Entry by: _____ Date: 02/09/99

Approved by: _____ Date: _____

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9556	
Work List	Tare Weight (W1)	1.8072	
28185	Weight of Solution (W2-W1)	0.1484	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4840	NA
Matrix			
LIQUID			
Sample #			
S98T002957			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	$\vee$ RESULT $\vee$		
02/04/99			
Time	Specific Gravity =	1.484	
03:14 PM			

Data Entry by:

Date: 02/09/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUP	REPLICATE
DUP	Gross Weight (W2)	1.9441	
Work List	Tare Weight (W1)	1.8003	
28185	Weight of Solution (W2-W1)	0.1438	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4380	NA
Matrix			
LIQUID			
Sample #			
S98T002957			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99			
Time	Specific Gravity =	1.438	
03:14 PM			

Data Entry by:	Date:	02/09/99
Approved by:	Date:	

LABCORE Completed Worklist Report for Worklist# 28186

Analyst: rwk

Instrument: BA001

Book#: 134N16C

Method: LA-510-112 Rev/Mod _____

Worklist Comment: BY105 FOR SPG-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		SPG-01	LIQUID	1.3779	1.384	98.991 % Recovery	
2 SAMPLE	S98T002958	0	SPG-01	LIQUID	N/A	1.465	1.00e-003	Sp.G.
3 DUP	S98T002958	0	SPG-01	LIQUID	1.465	1.447	1.236 RPD	
4 SAMPLE	S98T002959	0	SPG-01	LIQUID	N/A	1.455	1.00e-003	Sp.G.
5 DUP	S98T002959	0	SPG-01	LIQUID	1.455	1.435	1.384 RPD	
6 SAMPLE	S98T003002	0	SPG-01	LIQUID	N/A	1.436	1.00e-003	Sp.G.
7 DUP	S98T003002	0	SPG-01	LIQUID	1.436	1.437	0.070 RPD	
8 SAMPLE	S98T003003	0	SPG-01	LIQUID	N/A	1.447	1.00e-003	Sp.G.
9 DUP	S98T003003	0	SPG-01	LIQUID	1.447	1.438	0.624 RPD	

Final page for worklist# 28186

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 28186

Analyst: ML Instrument: BA001 Book # 134716-CMethod: LA-510-112 Rev/Mod E-0

Worklist Comment: BY105 FOR SPG-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID			N/A	Sp.G.
98000509	BY-105 (2)	2 SAMPLE	S98T002958	0	SPG-01	LIQUID	N/A			Sp.G.
98000509	BY-105 (2)	3 DUP	S98T002958	0	SPG-01	LIQUID			N/A	Sp.G.
98000509	BY-105 (2)	4 SAMPLE	S98T002959	0	SPG-01	LIQUID	N/A			Sp.G.
98000509	BY-105 (2)	5 DUP	S98T002959	0	SPG-01	LIQUID			N/A	Sp.G.
98000510	BY-105 (2)	6 SAMPLE	S98T003002	0	SPG-01	LIQUID	N/A			Sp.G.
98000510	BY-105 (2)	7 DUP	S98T003002	0	SPG-01	LIQUID			N/A	Sp.G.
98000510	BY-105 (2)	8 SAMPLE	S98T003003	0	SPG-01	LIQUID	N/A			Sp.G.
98000510	BY-105 (2)	9 DUP	S98T003003	0	SPG-01	LIQUID			N/A	Sp.G.

Final page for worklist # 28186

ML 2/4/99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		STANDARD	STANDARD
STANDARD	Gross Weight (W2)	2.0118	1.9221
Work List	Tare Weight (W1)	1.8748	1.7862
28186	Weight of Solution (W2-W1)	0.137	0.1359
Test Code	Volume of Solution μ L	100.0000	100.0000
SPG-01	Specific Gravity	1.3700	1.3590
Matrix	Specific Gravity (Average)	1.3645	
LIQUID			
Sample #			
STANDARD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity Average =	1.364	
03:30 PM			

Data Entry by:

Date: 02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	2.0032	0.0000
Work List	Tare Weight (W1)	1.8567	0.0000
28186	Weight of Solution (W2-W1)	0.1465	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4650	NA
Matrix			
LIQUID			
Sample #			
S98T002958			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L}/\text{mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.465	
03:30 PM			

Data Entry by:

Date: 02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	2.0043	0.0000
Work List	Tare Weight (W1)	1.8596	0.0000
28186	Weight of Solution (W2-W1)	0.1447	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4470	NA
Matrix			
LIQUID			
Sample #			
S98T002958			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BAOO1	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	$\vee$ RESULT $\vee$		
Time	Specific Gravity =	1.447	
03:30 PM			

Data Entry by:		Date:	02/10/99
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9936	0.0000
Work List	Tare Weight (W1)	1.8481	0.0000
28186	Weight of Solution (W2-W1)	0.1455	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4550	NA
Matrix			
LIQUID			
Sample #			
S98T002959			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BAOO1	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.455	
03:30 PM			

Data Entry by:	Date:	02/10/99
Approved by: <i>NA</i>	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.9596	0.0000
Work List	Tare Weight (W1)	1.8161	0.0000
28186	Weight of Solution (W2-W1)	0.1435	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4350	NA
Matrix			
LIQUID			
Sample #			
S98T002959			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.435	
03:30 PM			

Data Entry by:

Date:

02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9556	0.0000
Work List	Tare Weight (W1)	1.8120	0.0000
28186	Weight of Solution (W2-W1)	0.1436	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4360	NA
Matrix			
LIQUID			
Sample #			
S98T003002			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	ν RESULT ν		
Time	Specific Gravity =	1.436	
03:30 PM			

Data Entry by:

Date: 02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.9992	0.0000
Work List	Tare Weight (W1)	1.8555	0.0000
28186	Weight of Solution (W2-W1)	0.1437	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4370	NA
Matrix			
LIQUID			
Sample #			
S98T003002			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.437	
03:30 PM			

Data Entry by:	Date: 02/10/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.9651	0.0000
Work List	Tare Weight (W1)	1.8204	0.0000
28186	Weight of Solution (W2-W1)	0.1447	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4470	NA
Matrix			
LIQUID			
Sample #			
S98T003003			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BAOO1	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/04/99	v RESULT v		
Time	Specific Gravity =	1.447	
03:30 PM			

Data Entry by:

Date: 02/10/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	2.0111	0.0000
Work List	Tare Weight (W1)	1.8673	0.0000
28186	Weight of Solution (W2-W1)	0.1438	0
Test Code	Volume of Solution μ L	100.0000	0.0000
SPG-01	Specific Gravity	1.4380	NA
Matrix			
LIQUID			
Sample #			
S98T003003			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BAOO1	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
RWK	Specific Gravity = $[(W2-W1) \cdot 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} \cdot 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/04/99			
Time	Specific Gravity =	1.438	
03:30 PM			

Data Entry by:

Date:

02/10/99

Approved by:

Date:

LABCORE Completed Worklist Report for Worklist# 28187

Analyst: krm

Instrument: BA001

Book#: 134N16C

Method: LA-510-112 Rev/Mod _____

Worklist Comment: BY105 FOR SPG-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	SPG-01	LIQUID	1.3779	1.366	99.136 % Recovery	
2 SAMPLE	S98T003004 0	SPG-01	LIQUID	N/A	1.414	1.00e-003	Sp.G.
3 DUP	S98T003004 0	SPG-01	LIQUID	1.414	1.452	2.652 RPD	
4 SAMPLE	S98T003005 0	SPG-01	LIQUID	N/A	1.454	1.00e-003	Sp.G.
5 DUP	S98T003005 0	SPG-01	LIQUID	1.454	1.404	3.499 RPD	
6 SAMPLE	S98T003384 0	SPG-01	LIQUID	N/A	1.259	1.00e-003	Sp.G.
7 DUP	S98T003384 0	SPG-01	LIQUID	1.259	1.404	10.890 RPD	

Final page for worklist# 28187

Analyst Signature _____ Date _____

Mary Tracy
Analyst Signature _____ Date 2-12-99

[Signature]
Reviewer Signature _____ Date 2/16/99

LABCORE Data Entry Template for Worklist# 28187

Analyst: KRM Instrument: BA001 _____ Book # 134N16C

Method: LA-510-112 Rev/Mod E-0

Worklist Comment: BY105 FOR SPG-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			SPG-01	LIQUID	_____	_____	N/A	Sp.G.
98000510	BY-105 (2)	2 SAMPLE	S98T003004	0	SPG-01	LIQUID	N/A	_____	_____	Sp.G.
98000510	BY-105 (2)	3 DUP	S98T003004	0	SPG-01	LIQUID	_____	_____	N/A	Sp.G.
98000510	BY-105 (2)	4 SAMPLE	S98T003005	0	SPG-01	LIQUID	N/A	_____	_____	Sp.G.
98000510	BY-105 (2)	5 DUP	S98T003005	0	SPG-01	LIQUID	_____	_____	N/A	Sp.G.
98000437	BY-105 (2)	6 SAMPLE	S98T003384	0	SPG-01	LIQUID	N/A	_____	_____	Sp.G.
98000437	BY-105 (2)	7 DUP	S98T003384	0	SPG-01	LIQUID	_____	_____	N/A	Sp.G.

Final page for worklist # 28187

Analyst Signature Date 2-6-99

Analyst Signature Date 2-12-99

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		STD	REPLICATE
STD	Gross Weight (W2)	1.4633	1.5089
Work List	Tare Weight (W1)	1.3268	1.3721
28187	Weight of Solution (W2-W1)	0.1365	0.1368
Test Code	Volume of Solution μ L	100.0000	100.0000
SPG-01	Specific Gravity	1.3650	1.3680
Matrix	Specific Gravity (Average)	1.3665	
LIQUID			
Sample #			
STD			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/06/99	v RESULT v		
Time	Specific Gravity Average =	1.366	
12:00 AM			

Data Entry by:

Date: 02/12/99

Approved by:

Date:

Form 510112L1 Rev. 1.1

Page 1 of 1

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4552	
Work List	Tare Weight (W1)	1.3138	
28187	Weight of Solution (W2-W1)	0.1414	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4140	NA
Matrix			
LIQUID			
Sample #			
S98T003004			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.414	
12:00 PM			

Data Entry by:		Date:	02/12/99
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.4921	
Work List	Tare Weight (W1)	1.3469	
28187	Weight of Solution (W2-W1)	0.1452	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4520	NA
Matrix			
LIQUID			
Sample #			
S98T003004			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/06/99	$\vee$ RESULT $\vee$		
Time	Specific Gravity =	1.452	
12:00 AM			

Data Entry by:	Date: 02/12/99
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4519	
Work List	Tare Weight (W1)	1.3065	
28187	Weight of Solution (W2-W1)	0.1454	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4540	NA
Matrix			
LIQUID			
Sample #			
S98T003005			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.454	
12:00 AM			

Data Entry by:

Date: 02/12/99

Approved by:

Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.4492	
Work List	Tare Weight (W1)	1.3088	
28187	Weight of Solution (W2-W1)	0.1404	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4040	NA
Matrix			
LIQUID			
Sample #			
S98T003005			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.404	
12:00 AM			

Data Entry by:	Date: 02/12/99
Approved by:	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		SAMPLE	REPLICATE
SAMPLE	Gross Weight (W2)	1.4441	
Work List	Tare Weight (W1)	1.3182	
28187	Weight of Solution (W2-W1)	0.1259	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.2590	NA
Matrix			
LIQUID			
Sample #			
S98T003384			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date			
02/06/99	v RESULT v		
Time	Specific Gravity =	1.259	
12:00 AM			

Data Entry by:		Date:	02/12/99
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

SPECIFIC GRAVITY : LA-510-112 (E-0)

Type		DUPLICATE	REPLICATE
DUPLICATE	Gross Weight (W2)	1.5094	
Work List	Tare Weight (W1)	1.3690	
28187	Weight of Solution (W2-W1)	0.1404	0
Test Code	Volume of Solution μ L	100.0000	
SPG-01	Specific Gravity	1.4040	NA
Matrix			
LIQUID			
Sample #			
S98T003384			
Instrument Code	Gross Weight (W2) = Wt. of vial + cap + cotton + solution		
BA001	Tare Weight (W1) = Wt. of vial + cap + cotton		
Analyst			
KRM	Specific Gravity = $[(W2-W1) * 1000 \mu\text{L/mL}] / [\text{Vol. of Solution } \mu\text{L} * 1.000 \text{ g/mL}]$		
Date	v RESULT v		
02/06/99			
Time	Specific Gravity =	1.404	
12:00 AM			

Data Entry by:

Date: 02/12/99

Approved by:

Date:

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 1 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 1 THRU 347)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	NARRATIVE, CORRESPONDENCE, SAMPLE DATA SUMMARY, CHAIN OF CUSTODY FORMS, PHOTOGRAPHS, SAMPLE HANDLING, CORE COMPOSITE WORKSHEETS, EXTRUSION WORKLISTS, SAMPLE PREPARATIONS, FUSION & WATER & ACID DIGESTION WORKSHEETS & BULK DENSITY WORKSHEETS		

APR 06 1999

ENGINEERING DATA TRANSMITTAL

Page 1 of 1
1. EDT 626135

2. To: (Receiving Organization) Distribution		3. From: (Originating Organization) Waste Management Laboratory		4. Related EDT No.: N/A	
5. Proj./Prog./Dept./Div.: Tank 241-BY-105 Core/Waste Management/PPC/WML		6. Design Authority/ Design Agent/Cog. Engr.: Franciska H. Steen		7. Purchase Order No.: N/A	
8. Originator Remarks: This document is being released into the supporting document system for retrievability purposes.				9. Equip./Component No.: N/A	
				10. System/Bldg./Facility: 241-BY-105	
11. Receiver Remarks: For release.		11A. Design Baseline Document? ☐ Yes ☒ No		12. Major Assm. Dwg. No.: N/A	
				13. Permit/Permit Application No.: N/A	
				14. Required Response Date: 04/08/99	

15. DATA TRANSMITTED					(F)	(G)	(H)	(I)
(A) Item No.	(B) Document/Drawing No.	(C) Sheet No.	(D) Rev. No.	(E) Title or Description of Data Transmitted	Approval Designator	Reason for Transmittal	Originator Disposition	Receiver Disposition
1	HNF-1663	N/A	0	Tank 241-BY-105, Cores 246, 246R, 250, 251 and 252 Analytical results for the Final Report	Q	2	1	1

16. KEY		
Approval Designator (F)	Reason for Transmittal (G)	Disposition (H) & (I)
E, S, Q, D or N/A (see WMC-CM-3-5, Sec.12.7)	1. Approval 4. Review 2. Release 5. Post-Review 3. Information 6. Dist. (Receipt Acknow. Required)	1. Approved 4. Reviewed no/comment 2. Approved w/comment 5. Reviewed w/comment 3. Disapproved w/comment 6. Receipt acknowledge

17. SIGNATURE/DISTRIBUTION (See Approval Designator for required signatures)									
(G) Reason	(H) Disposition	(J) Name	(K) Signature (M) MSIN	(L) Date	(G) Reason	(H) Disposition	(J) Name	(K) Signature (M) MSIN	
		Design Authority							
		Design Agent							
2	1	Cog. Eng. F. H. Steen	<i>[Signature]</i>	4/5/99					
2	1	Cog. Mgr. D. B. Hardy	<i>[Signature]</i>	4/5/99					
2	1	QA L. P. Markel	<i>[Signature]</i>	4-5-99					
		Safety							
		Env.							
18. L. A. Diaz Signature of EDT Date 4/5/99 Originator			19. N/A Authorized Representative Date for Receiving Organization			20. D. B. Hardy Signature of Design Authority Date 4/5/99 Cognizant Manager			
21. DOE APPROVAL (if required) Ctrl. No. ☐ Approved ☐ Approved w/comments ☐ Disapproved w/comments									

DISTRIBUTION SHEET

To Distribution	From Production Planning & Control	Page 1 of 1		
		Date: 04/08/99		
Project Title/Work Order HNF-1663 Rev. 0 "Tank 241-BY-105, Cores 246, 246R, 250, 251 and 252 Analytical Results for the Final Report"		EDT NO.: EDT-626135		
		ECN NO.: N/A		
Name	MSIN	Text With all Attach	EDT/ECN ONLY	

U.S. Department of Energy, RL
C. A. Babel

S7-54 X

Lockheed Martin Hanford Corp.

J. N. Appel	A3-03			X
R. A. Bechtold	K7-22	X		
T. M. Blaak	S5-13	X		
T. M. Brown	S5-05			X
K. G. Carothers	R2-11			X
J. G. Field	R2-12			X
J. Jo	R2-12			X
M. J. Kupfer	R3-75			X
R. E. Larson	T4-07			X
L. M. Sasaki	R2-12			X
J. B. Schaffer	R2-12			X
L. A. Stauffer	R2-11			X
J. A. Voogd	R2-53			X
A. E. Young	R1-10	X		
B. C. Simpson	R2-12	X		

Lockheed Martin Services, Inc.

Central Files B1-07 X

Pacific Northwest National Laboratory

K. L. Silvers K9-08 X

Waste Management of Hanford, Inc.

B. V. Burrows	H6-30			X
L. A. Diaz	T6-12			X*
D. B. Hardy	T6-12			X
D. C. Hetzer	S6-31			X
J. E. Hyatt	T6-14			X
K. L. Powell	S3-30			X
C. M. Seidel	S3-30			X
F. H. Steen	T6-12			X
LTIC	T6-03			X

SGN Eurisys Services Corp.

B. A. Higley	R3-73			X
W. S. Callaway	S3-90			X

Defense Nuclear Federal Safety Board

Rich Tontodonato X
625 Indiana Ave. N. W.
Washington D.C. 20004

DOE/RL Reading Room H2-53 X

* Needs only releasing paperwork, not a copy of the released document.

S

Tank 241-BY-105, Cores 246, 246R, 250, 251 and 252 Analytical Results for the Final Report

Franciska H. Steen

Waste Management of Hanford, Inc., Richland, WA 99352

U.S. Department of Energy Contract DE-AC06-96RL13200

EDT/ECN: EDT-626135

UC: 2070

Org Code: 31B00

Charge Code: LC087

B&R Code: EW 3120074

Total Pages: ~~5114~~ *5114-5118*

5128 7/5/99

Key Words: Tank-241-BY-105, B Farm, BY-105, 241-BY-105 Tank Characterization, Core, Analytical Results, Final Report

Abstract: N/A

TRADEMARK DISCLAIMER. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

Printed in the United States of America. To obtain copies of this document, contact: Document Control Services, P.O. Box 950, Mailstop H6-08, Richland WA 99352, Phone (509) 372-2420; Fax (509) 376-4989.

Franciska H. Steen 4/4/99
Release Approval Date

DATE:		ID:
STA: <i>A</i>		<i>2.</i>
APR 06 1999		

Release Stamp

Approved for Public Release

HNF-1663 REV. 0

WASTE MANAGEMENT LABORATORY

**TANK 241-BY-105, CORES 246, 246R, 250, 251
AND 252
ANALYTICAL RESULTS FOR THE FINAL REPORT**

Project Coordinator: Franciska H. Steen

**Prepared for the U.S. Department of Energy
Office of Environmental Restoration
and Waste Management**

by

**222-S Laboratory
Waste Management of Hanford, Inc.
P.O. Box 700
Richland, Washington**

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955, 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091
Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

Sec 17
of 17
Continued

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

HNF-1663 REV. 0

NARRATIVE

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

**WASTE MANAGEMENT LABORATORY
TANK 241-BY-105, CORES 246, 246R, 250, 251 AND 252,
ANALYTICAL RESULTS FOR THE FINAL REPORT**

This document is the analytical laboratory report for tank 241-BY-105 rotary mode core segments collected between July 23, 1998 and September 4, 1998. The segments were subsampled and analyzed in accordance with the *Tank 241-BY-105 Rotary Mode Core Sampling and Analysis Plan* (TSAP) (Simpson, 1998), the *Safety Screening Data Quality Objective* (DQO) (Dukelow, et al., 1995) and the *Flammable Gas Data Quality Objective* (DQO) (Bauer, et al., 1998). The analytical results are included in the data summary table (Table 1).

None of the samples submitted for Differential Scanning Calorimetry (DSC), Total Alpha Activity (AT) and Total Organic Carbon (TOC) exceeded notification limits as stated in the TSAP (Simpson, 1998). The statistical results of the 95% confidence interval on the mean calculations are provided by the Tank Waste Remediation Systems Technical Basis Group in accordance with the Memorandum of Understanding (Schreiber, 1997) and not considered in this report.

The TSAP (Simpson, 1998) required viscosity, shear strength, particle size distribution and settling behavior analyses be performed on the lower half portion of each segment. None of these analyses were performed for the reasons discussed below.

The flammable gas program rescinded the requirement for the settling behavior study (Attachment 2). There was insufficient sample material to perform both the shear strength and viscosity analyses. The particle size analysis was requested on the stratum composites. However, due to the homogenization and filtering of these materials, the chemist determined the results of the particle size analysis would be unusable. Therefore, this analysis was not performed.

Appearance and Sample Handling

Attachment 1 is a cross reference to relate the tank farm identification numbers to the 222-S Laboratory LabCore/LIMS sample numbers. The subsamples generated in the laboratory for analyses are identified in these diagrams with their sources shown.

Two core samples, each consisting of nine segments, were expected to be taken from tank 241-BY-105. However, due to poor sample recovery, five cores were taken and identified as 246, 246R, 250, 251 and 252.

Core 246

Seven rotary mode core segments (1, 2, 3, 4, 8, 9 and 9A) were removed from tank 241-BY-105 riser 11B between July 23, 1998 and July 31, 1998. Segments were received by the 222-S Laboratory between August 7, 1998 and August 20, 1998.

HNF-1663, REV. 0

The TSAP (Simpson, 1998) states the core samples should be transported to the laboratory within three calendar days from the time each segment is removed from the tank, this requirement was not met for the segments from core 246. Table 2 summarizes the extrusion information.

Core 246R

Eight rotary mode core segments (2R, 3R, 4R, 5R, 6R, 7R, 8R and 9) were removed from tank 241-BY-105 riser 11B between August 3, 1998 and August 6, 1998. Segments were received by the 222-S Laboratory on August 20 and August 21, 1998.

The TSAP (Simpson, 1998) states the core samples should be transported to the laboratory within three calendar days from the time each segment is removed from the tank, this requirement was not met for the segments from core 246R. Table 3 summarizes the extrusion information.

Core 250

Ten rotary mode core segments (1, 2, 3, 4, 5, 6, 7, 8, 9A and 9B) were removed from tank 241-BY-105 riser 7 between August 18, 1998 and August 24, 1998. Segments were received by the 222-S Laboratory between August 21, 1998 and September 15, 1998.

The TSAP (Simpson, 1998) states the core samples should be transported to the laboratory within three calendar days from the time each segment is removed from the tank, this requirement was not met for the segments from core 250. Table 4 summarizes the extrusion information.

Core 251

Six rotary mode core segments (1, 2, 3, 4, 4A and 5) were removed from tank 241-BY-105 riser 7 between August 24, 1998 and August 31, 1998. Segments were received by the 222-S Laboratory on September 8, 1998.

The TSAP (Simpson, 1998) states the core samples should be transported to the laboratory within three calendar days from the time each segment is removed from the tank, this requirement was not met for the segments from core 251. Table 5 summarizes the extrusion information.

Core 252

Five rotary mode core segments (5, 6, 7, 8 and 9) were removed from tank 241-BY-105 riser 7 between September 2, 1998 and September 4, 1998. Segments were received by the 222-S Laboratory on September 9, 1998.

The TSAP (Simpson, 1998) states the core samples should be transported to the laboratory within three calendar days from the time each segment is removed from the tank, this requirement was not met for the segments from core 252. Table 6 summarizes the extrusion information.

HNF-1663, REV. 0

Field Blank

A field blank was provided to the 222-S laboratory with core 246. It underwent the same analysis as the drainable liquid as instructed by the TSAP (Simpson, 1998).

Hydrostatic Head Fluid

A sample of the hydrostatic head fluid (HHF) lithium bromide solution was provided with core 246 and was submitted to the laboratory for IC and ICP analyses.

Stratum Composites

Since there was insufficient material to form complete stratum composites for each core, the liquid and solid stratum composites were prepared per instruction from the customer (Attachment 2). The stratum composite worksheets are included in this document.

Table 2. Sample Receipt and Extrusion Information for 241-BY-105, Core 246.

Customer No.	Segment	Date Sampled	Date Received	Date Extruded	Water Extruded	Liquid Extrusion (%)	Solids Recovered (g)	Sample Description
246-01	1	07/27/98	08/07/98	08/16/98	4.0	0.0	68.8-Lower half	The solids were yellow-white in color and resembled a dry salt.
246-02	2	07/27/98	08/07/98	08/16/98	8.0	0.0	204.4-Lower half	The solids were yellow-white in color and resembled a dry salt.
246-03	3	07/27/98	08/07/98	08/16/98	1.0	0.0	14.6-Lower half	The solids were light gray in color and resembled a dry salt.
246-04	4	07/27/98	08/07/98	08/16/98	3.0	0.0	104.1-Lower half	The solids were white and gray in color and the texture resembled a dry salt.
246-08	8	07/20/98	08/20/98	08/31/98	2.0	0.0	42.6-Lower half	The solids were brown and white in color and resembled a moist salt.
246-09	9	07/31/98	08/07/98	08/16/98	4.0	0.0	139.7-Lower half	The solids were medium brown in color and resembled a wet sludge.
246-9A	9A	07/31/98	08/07/98	08/16/98	1.0	0.0	42.3-Lower half	The solids were medium brown in color and resembled a wet sludge.

Table 3. Sample Receipt and Extrusion Information for 241-BY-105, Core 246R.

Container ID	Segment	Date Sampled	Date Received	Date Extruded	Isotopes Provided	Liquid Recovered (g)	Solids Recovered (g)	Sample Description
242R-02R	2R	08/03/98	08/21/98	09/01/98	0.0	356.6-Drainable	0.0	The liquid was light brown in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
242R-03R	3R	08/04/98	08/20/98	09/01/98	0.0	363.4-Drainable	0.0	The liquid was light brown in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
242R-04R	4R	08/04/98	08/20/98	09/01/98	0.0	334.2-Drainable	0.0	The liquid was light brown in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
242R-05R	5R	08/04/98	08/20/98	08/31/98	0.0	114.8-Drainable	0.0	The liquid was light brown in color and opaque. Extruded 300 mL of liquid. Retained 65 mL of liquid. No organic layer was observed. Extruded a small amount of solids which were added to the liquid.
242R-06R	6R	08/05/98	08/21/98	08/31/98	0.0	374.7-Drainable	0.0	The liquid was light brown in color and opaque. Collected 300 mL of liquid. No organic layer was observed. Extruded a small amount of solids which were added to the liquid.
242R-07R	7R	08/06/98	08/20/98	09/01/98	0.0	354.2-Drainable	0.0	The liquid was light brown in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
242R-08R	8R	08/06/98	08/20/98	09/01/98	2.0	348.5-Drainable	92.1-Lower half	The solids were brown in color and resembled a salt slurry. The liquid was light brown in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
242R-09	9	08/06/98	08/21/98	09/01/98	8.0	151.2-Drainable	275.0-Lower half	The solids were brown in color and resembled a salt slurry. The liquid was brown in color and opaque. Collected 100 mL of liquid. No organic layer was observed.

Table 4. Sample Receipt and Extrusion Information for 241-BY-105, Core 250.

Clipboard #	Segment	Date Sampled	Date Received	Date Extruded	Index Provided	Liquid Recovered (%)	Solids Recovered (g)	Sample Description
250-01	1	08/18/98	08/21/98	09/02/98	0.0	0.0	0.0	Sampler empty.
250-02	2	08/18/98	08/21/98	09/02/98	0.0	0.0	0.0	Sampler empty.
250-03	3	08/18/98	08/21/98	09/02/98	2.0	0.0	92.9-Lower half	The solids were dirty white in color and resembled a dry salt.
250-04	4	08/19/98	09/03/98	09/14/98	3.0	0.0	58.7-Lower half	The solids were brown and white in color and resembled a dry salt.
250-05	5	08/19/98	09/03/98	09/14/98	5.0	0.0	122.6-Lower half	The solids were brown and white in color and resembled a dry salt.
250-06	6	08/20/98	09/03/98	09/14/98	7.0	0.0	157.2-Lower half	The solids were gray and white in color and resembled a dry salt.
250-07	7	08/20/98	09/03/98	09/18/98	4.0	0.0	60.0-Lower half	The solids were gray and white in color and resembled a dry salt. Extruded a small amount of liquid which was added to the solids.
250-08	8	08/20/98	09/15/98	09/27/98	0.0	0.0	0.0	Sampler empty.
250-09	9A	08/21/98	09/03/98	09/18/98	0.0	0.0	0.0	Sampler empty.
250-9B	9B	08/24/98	09/03/98	09/18/98	2.0	210.2-Drainable	71.1-Upper half 33.0-Lower half	The upper 0.5 inches of solids were white in color and resembled a moist salt. The next 1.5 inches were brown in color and resembled a wet sludge. The liquid was yellow/brown in color and opaque. Collected 160 mL of liquid. No organic layer was observed.

Table 5. Sample Receipt and Extrusion Information for 241-BY-105, Core 251.

Customer ID	Sample ID	Date Sampled	Date Received	Date Released	Isotopes Found	Liquid Recovered (g)	Solids Recovered (g)	Sample Description
251-1	1	08/24/98	09/08/98	09/18/98	0.0	357.0-Drainable	0.0	The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
251-2	2	08/24/98	09/08/98	09/18/98	3.0	334.4-Drainable	68.4-Lower half	The solids were white, brown and green in color and resembled a wet salt. The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
251-3	3	08/24/98	09/08/98	09/18/98	0.0	365.1-Drainable	0.0	The liquid was yellow-green in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
251-4	4	08/25/98	09/08/98	09/27/98	4.0	330.8-Drainable	103.6-Lower half	The solids were black and white in color and resembled a dry salt. The liquid was amber in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
251-04A	4A	08/27/98	09/08/98	09/27/98	3.0	0.0	31.4-Lower half	The solids were black and white in color and resembled a dry salt.
251-05	5	08/31/98	09/08/98	09/27/98	8.0	128.8-Drainable	289.1-Lower half	The solids were black and white in color and resembled a moist salt. The liquid was amber in color and opaque. Collected 100 mL of liquid. No organic layer was observed.

Table 6. Sample Receipt and Extrusion Information for 241-BY-105, Core 252.

Container ID	Segment	Date Sampled	Date Extruded	Index Extruded	Liquid Recovered (g)	Solid Recovered (g)	Sample Description
252-05	5	09/02/98	09/09/98	0.0	326.0-Drainable	0.0	The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
252-06	6	09/02/98	09/09/98	0.0	326.0-Drainable	0.0	The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
252-07	7	09/03/98	09/09/98	0.0	361.4-Drainable	0.0	The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
252-08	8	09/03/98	09/09/98	0.0	386.2-Drainable	0.0	The liquid was yellow in color and opaque. Collected 250 mL of liquid. No organic layer was observed.
252-9	9	09/04/98	09/27/98	<1.0	0.0	0.0	Less than one inch of solids was extruded. Did not retain. Insufficient amount of material for analyses.

Analytical Results Summary

The data summary report (Table 1) compiles the analytical results that satisfy all applicable DQOs.

As noted in the extrusion information in Tables 2, 3, 4, 5 and 6, the solid material collected from Tank 241-BY-105 was primarily a dry salt and a wet sludge. There were various colors of material present gray, brown, yellow and white. It was impossible to separate the different colored material. Obtaining reproducible results with this non-homogenous material was very difficult and may have been the main reason for the large difference between the sample, duplicate and triplicate (were applicable) measurements for many of the analyses. In many cases reanalysis did not improve the results.

Inorganic Analyses

Differential Scanning Calorimetry (DSC)

The DSC analyses were performed in duplicate on direct subsamples. The exothermic energy based on the dry weight of the subsample was calculated for all subsamples. The average of the TGA results for each subsample was used in the dry weight correction for that subsample. The results for sixteen samples were the sum of two or more exotherms. More information may be obtained by examining the raw data.

Relative percent differences (RPDs) outside of the required range (30% for solids, 20% for liquids) were reported for nine of the forty-three subsamples submitted for analysis. The nonhomogeneous material and small sample size required for this analysis made it difficult to obtain reproducible results. The chemist noted the high RPDs were due to sample inhomogeneity and no reanalyses were performed. However, two samples had triplicate analyses performed. The results are presented in Table 7.

Table 7. Triplicate Results for DSC Analysis

Labcore ID	Sample Result Joules/gram	Duplicate Result Joules/gram	Triplicate Result Joules/gram
S98T002957	4.99	15.89	14.62
S98T002981	65.65	17.96	16.01

For the samples that did not exhibit exothermic energy, the calculation for the RPD is indeterminate because of the division by zero. Due to the way the database performs calculations, zero divided by zero is reported as "0.0". The standard recoveries for this analysis were within the required limits of 80%-120%.

Thermogravimetric Analysis (TGA)

The TGA analyses were performed in duplicate on direct subsamples. Typically TGA results are determined by summing the weight loss steps which occurred below 250°C; weight loss steps above this are not used to determine the result. However, for tank 241-BY-105 approximately 30% of the thermograms showed weight loss beyond 250°C. The results for fourteen of the forty-three subsamples were the sum of two or more weight loss steps. More information may be obtained by examining the raw data.

RPDs outside of the required range (30% for solids, 20% for liquids) were reported for four of the forty-three subsamples submitted for analysis. The nonhomogeneous material and small sample size required for this analysis made it difficult to obtain reproducible results. The chemist noted the high RPDs were due to sample inhomogeneity. The chemist requested triplicate analyses be performed on two subsamples. The results are presented in Table 8.

Table 8. Triplicate Results for TGA Analysis

Labcore ID	Sample Result	Duplicate Result	Triplicate Result
S98T002910	5.36	10.21	9.13
S98T002911	2.99	4.01	4.32

The standard recoveries for this analysis were within the required limits of 80%-120%.

Bulk Density

Bulk density was performed on the solid subsamples as required by the TSAP (Simpson, 1998), the exceptions are noted below. The results of the bulk density test ranged from 1.74 g/mL to 2.18 g/mL. The bulk density results were greater than the 1.5 g/mL used to determine the solid total alpha activity notification limit stated in the TSAP (Simpson, 1998). Therefore the bulk density result of 2.18 g/mL was used to recalculate the solid total alpha notification limit for the tank.

There was insufficient sample material to perform the bulk density analysis on segment 3 from core 246. The bulk density analysis was attempted for segments 1, 2, 4 and 8 from core 246, segments 3, 4, 5, 6 and 9B from core 250 and segments 2 and 4A from core 251. These samples were very dry and accurate volumes could not be measured due to the void spaces in the material. No results were obtained as indicated by the "dry" in the results column of Table 1.

In addition, the bulk density analysis was attempted for segments 9 and 9A from core 246. These samples adhered to the side of the container used during the bulk density analysis and accurate volumes could not be measured. No results were obtained as indicated by the "adhesive" in the results column of Table 1. There were no quality control (QC) parameters stated in the TSAP (Simpson, 1998) for these samples.

Specific Gravity (SpG)

Specific gravity analyses were performed in duplicate on the direct liquid subsamples. The results for this analysis are provided in Table 1. There was insufficient sample to perform this analysis on the field blank. The RPDs were less than 10%. The standard recoveries were within the control limits of the laboratory.

Hydroxide (OH)

The OH analysis was performed on direct subsamples of liquid as indicated by a blank in the A# column in Table 1. The solid subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1.

High RPDs (>10%) were reported for two of the twenty-three subsamples submitted for analysis. The sample results were near to or just above ten times the detection limit which decreased the precision of the analysis making it difficult to obtain the precision requirement of less than 10% as stated in the TSAP (Simpson, 1998). The hydroxide concentration was consistent with the other samples and no reanalysis was requested. The standard recoveries were within the required limits of 90% - 110%.

Ammonia (NH₃)

The NH₃ analysis was performed in duplicate on direct liquid subsamples as indicated by a blank in the A# column in Table 1.

High RPDs (>10%) were reported for nine of the twenty subsamples submitted for analysis. The concentration of ammonia in the sample was less than ten times the detection limit which decreased the precision of the analysis. No reanalysis was requested. However, selected samples had triplicate analysis performed. The results are presented in Table 9.

Table 9. Triplicate Results for Ammonia Analysis

Labcore ID	Sample Result $\mu\text{g/mL}$	Duplicate Result $\mu\text{g/mL}$	Triplicate Result $\mu\text{g/mL}$
S98T002806	2.76E+24.99	1.78E+2	1.89E+2
S98T002807	1.86E+2	6.02E+1	4.11E+2

HNF-1663, REV. 0

Standard recoveries outside of the required range (90% - 110%) were reported for five of the twenty subsamples submitted for analysis. The standard recoveries were within the method control limits statistically determined from historical data and no reanalyses were requested. The spike recoveries were within the required limits of 75%-125%.

Chromium VI (CrVI)

Chromium VI analyses was performed in duplicate direct liquid subsamples. Solid subsamples were prepared for analysis by performing a water digest in duplicate. The water digest is indicated with an "W" in the A# column in Table 1.

The RPDs were less than 20%, the standard recoveries were within the required limits of 80%-120% and the spike recoveries were within the required limits of 75%-125%.

Ion Chromatography (IC)

The IC analyses were performed on direct subsamples of liquid as indicated by a blank in the aliquot class (A#) column in Table 1. The solid subsamples were prepared for analysis by performing a water digest. This is indicated with a "W" in the A# column in Table 1.

The required analytes were bromide (Br^-), fluoride (F^-), chloride (Cl^-), nitrate (NO_3^-), nitrite (NO_2^-), phosphate (PO_4^{3-}), sulfate (SO_4^{2-}), oxalate and formate. All other analyte results are presented in Appendix A. These results are considered "opportunistic" and do not have customer defined QC parameters. Therefore, any anomalies in those results are not discussed in this report. There was insufficient sample to perform the IC-anions on the drainable liquid portion of core 250 segment 9B. However, the IC-organic acid analysis was performed.

High RPDs ($>20\%$) were reported for a small percentage of samples analyzed for the following: Cl^- , F^- , formate, NO_2^- , NO_3^- , oxalate, PO_4^{3-} and SO_4^{2-} . Selected samples were reanalyzed with little or no improvement in the RPDs. The high RPDs were attributed to sample inhomogeneity and no further reanalyses were requested. More information may be obtained by examining the raw data.

HNF-1663, REV. 0

Spike recoveries outside of the required range (75%-125%) were reported for a small percentage of samples analyzed for the following: Br⁻, Cl⁻, F⁻, formate, NO₃⁻ and NO₂⁻, oxalate, PO₄⁻³ and SO₄⁻². The spike failures for Cl⁻, F⁻, NO₃⁻, NO₂⁻, oxalate, PO₄⁻³ and SO₄⁻² were attributed to the high concentration of these analytes in the samples with respect to the amount of spike standard added. The spike failures for F⁻ were also attributed to matrix interferences from small organic acids. The spike failure for formate was attributed to matrix interferences from metals and since these interfering analytes could not be removed from the matrix, no reanalysis was requested. The standard recoveries were within the required limits of 80%-120%.

Inductively Coupled Plasma Spectrophotometry (ICP)

The liquid subsamples were prepared for analysis by an acid adjustment of the direct subsample. This is indicated by a "D" in the A# column in Table 1. The solid subsamples were prepared for analysis by performing both an acid digest as indicated by an "A" and a fusion as indicated by an "F" in the A# column in Table 1.

The required analytes were aluminum (Al), boron (B), barium (Ba), bismuth (Bi), calcium (Ca), chromium (Cr), iron (Fe), lithium (Li), manganese (Mn), sodium (Na), nickel (Ni), potassium (K), silicon (Si), uranium (U), zinc (Zn) and zirconium (Zr). All other analyte results are presented in Appendix A. These results are considered "opportunistic" and do not have customer defined QC parameters. Therefore, any anomalies in those results are not discussed in this report.

The fusion digestion process is performed using potassium hydroxide as a reagent and a nickel crucible for the digestion vessel. Therefore, for the samples requiring fusion digestion, results for K are not available. The results for Ni are biased high due to leaching from the crucible. The results for K and more accurate results for Ni may be obtained from the acid digestion portion of the samples.

High RPDs (>20%) were reported for a small percentage of samples analyzed for the following: Al, Ca, Cr, Fe, Mn, K, U, Zn and Zr. It should be noted that most of the failures occurred on the acid digested sample aliquots. In certain matrices, some metals may not have been completely or uniformly dissolved by the acid digestion process. However, the concentration of many of these analytes were less than ten times the detection limit and the precision of the analysis was decreased. No reparation or reanalysis was requested.

High RPDs (>20%) were reported for approximately 50% of the samples analyzed for B, Ni and Si. These three analytes in particular had high RPDs due to non-uniform leaching from the digestion vessels used for the sample and duplicate aliquots. Since this leaching could not be avoided, no reanalyses were requested.

HNF-1663, REV. 0

Matrix spike recoveries outside of the 75% to 125% control limits as set by the TSAP (Simpson, 1998) were reported for a small percentage of samples analyzed for the following: Bi, Ca, Fe, Ni, K, U, Zn and Zr. In addition, spike failures were noted for approximately 50% of the samples analyzed for Al, Na and Si. The spike failures were attributed to the high concentration of these analytes in the samples with respect to the amount of spike standard added. A post digestion spike analysis was performed as an additional instrument performance check. The post digestion spike recoveries were within the required limits. More information may be obtained by examining the raw data.

Typically, with high concentrations of analytes, it is difficult to add sufficient spike to perform a meaningful analysis. Therefore, the assessment of the accuracy of the measurement for these analytes were also made by comparison of the sample results to those of a serial dilution of the sample. The serial dilution was performed by preparing and analyzing an additional five-fold dilution of the sample. The result obtained from this analysis should be within $\pm 10\%$ of the undiluted sample result. The results of this comparison are presented in Table 10 and indicate the accuracy of this analysis was acceptable for all but one sample analyzed for Bi, five samples analyzed for K, six samples analyzed for Si and one sample analyzed for Zr. The undiluted results for these analytes were already near the detection limit and an additional five-fold dilution decreased the precision of the analysis even further. No reanalysis was requested.

The standard recoveries were within the required limits of 80%-120% for all analytes except for a small percentage of samples analyzed for Na, Si and Zn. The Si standard failures were due to the non-uniform leaching from the digestion vessels. Na and Zn standard failures were due to contamination found in the laboratory glassware, surgeon's gloves and plastic containers. Since these are common contaminants in the laboratory environment it was difficult to prevent this contamination. No reanalyses were requested.

Table 10: ICP Serial Dilution Results for Tank 241-BY-105

Sample ID	Analyte	Undiluted Sample Result (µg/mL)	Serial Dilution Result (µg/mL)	Percent Difference (%)
S98T002602	Na	2.82E+02	2.74E+02	2.8
S98T002604	Al	1.52E+01	1.48E+01	2.8
S98T002604	Na	1.48E+03	1.42E+03	4.0
S98T002608	Bi	1.54E+01	1.46E+01	5.6
S98T002608	Ca	1.55E+02	1.43E+02	7.9
S98T002608	Fe	1.48E+02	1.35E+02	8.7
S98T002608	Ni	6.49E+01	5.84E+01	9.9
S98T002608	Si	6.91E+00	6.52E+00	5.7
S98T002608	Na	6.30E+02	6.04E+02	4.2
S98T002608	U	1.92E+02	1.85E+02	3.4
S98T002608	Zn	6.78E-01	6.16E-01	9.2
S98T002608	Zr	2.24E-02	4.87E-02	117.3
S98T002739	Na	4.42E+02	4.32E+02	2.3
S98T002740	Al	2.61E+02	2.53E+02	2.8
S98T002740	K	7.31E+00	3.72E+00	49.0
S98T002740	Si	3.81E+00	2.55E+00	33.1
S98T002740	Na	1.09E+03	1.05E+03	3.5
S98T002819	Al	4.96E+04	4.85E+04	2.3
S98T002819	K	5.16E+03	4.85E+03	6.1
S98T002819	Na	2.24E+05	2.13E+05	4.7
S98T002844	Na	4.38E+02	4.20E+02	4.0
S98T002846	Al	7.80E+01	7.50E+01	3.8
S98T002846	K	1.15E+03	1.10E+03	4.7
S98T002846	Si	5.15E+00	4.03E+00	21.9
S98T002846	Na	1.15E+03	1.10E+03	4.7
S98T002920	Na	5.21E+02	4.95E+02	4.9
S98T002927	Al	2.54E+01	2.41E+01	5.2
S98T002927	K	8.21E-01	1.71E+00	108.6
S98T002927	Si	2.03E+00	1.65E+00	18.8
S98T002928	K	3.86E+00	1.92E+00	50.3
S98T002928	Si	2.07E+00	1.92E+00	7.3
S98T002928	Na	1.30E+03	1.24E+03	5.0

$$\% \text{ Difference} = (\text{ABS}(\text{Sample} - \text{Serial})) / \text{Sample} \times 100$$

Table 10: ICP Serial Dilution Results for Tank 241-BY-105

Sample ID.	Analyte	Undiluted Sample Result (µg/mL)	Serial Dilution Result (µg/mL)	Percent Difference (%)
S98T002930	Bi	4.13E+00	4.85E+00	17.5
S98T002930	Ca	1.41E+02	1.36E+02	3.1
S98T002930	Fe	5.50E+01	5.26E+01	4.3
S98T002930	Ni	2.80E+01	2.67E+01	4.7
S98T002930	Si	1.74E+01	1.68E+01	3.3
S98T002930	Na	1.19E+03	1.15E+03	3.4
S98T002930	U	8.31E+01	8.07E+01	2.9
S98T002939	Al	5.00E+04	4.92E+04	1.5
S98T002939	Na	2.23E+05	2.15E+05	3.6
S98T002960	Al	2.74E+04	2.67E+04	2.3
S98T002960	Na	2.18E+05	2.07E+05	4.7
S98T002962	Al	3.91E+04	3.84E+04	1.6
S98T002962	Na	2.19E+05	2.08E+05	5.3
S98T002963	Al	4.78E+04	4.72E+04	1.2
S98T002963	K	4.82E+03	4.79E+03	0.6
S98T002963	Na	2.17E+05	2.11E+05	3.1
S98T002990	Al	1.38E+02	1.35E+02	2.7
S98T002990	Ca	1.02E+01	1.00E+01	1.5
S98T002990	Fe	1.11E+01	1.08E+01	3.0
S98T002990	K	3.54E+00	7.35E+00	107.8
S98T002990	Si	5.96E-01	2.61E-01	56.3
S98T002990	Na	9.77E+02	9.31E+02	4.7
S98T002991	Al	1.89E+02	1.85E+02	2.1
S98T002991	Ca	7.88E+00	7.61E+00	3.4
S98T002991	Fe	7.88E+00	7.63E+00	3.3
S98T002991	Si	6.56E+00	6.36E+00	3.0
S98T002991	Na	8.88E+02	8.58E+02	3.4
S98T002993	Na	9.68E+02	9.63E+02	0.5
S98T003008	Al	4.49E+04	4.46E+04	0.7
S98T003008	Na	2.07E+05	2.03E+05	1.8
S98T003384	Al	3.91E+04	3.83E+04	2.1
S98T003384	Na	2.25E+05	2.14E+05	4.8

$$\% \text{ Difference} = (\text{ABS}(\text{Sample} - \text{Serial})) / \text{Sample} \times 100$$

Table 10: ICP Serial Dilution Results for Tank 241-BY-105

Sample ID	Analyte	Undiluted Sample Result ($\mu\text{g/mL}$)	Serial Dilution Result ($\mu\text{g/mL}$)	Percent Difference (%)
S98T003388	Na	5.08E+02	4.89E+02	3.7
S98T003389	Al	7.90E+01	7.63E+01	3.3
S98T003389	K	5.89E-01	2.49E+00	322.3
S98T003389	Si	1.51E+00	1.34E+00	11.0
S98T003389	Na	1.18E+03	1.12E+03	4.7
S98T003393	Na	5.17E+02	4.94E+02	4.5
S98T003394	Al	4.21E+01	4.01E+01	4.8
S98T003394	Fe	1.07E+01	1.01E+01	5.9
S98T003394	Si	1.82E+00	1.53E+00	15.8
S98T003394	Na	1.29E+03	1.22E+03	5.9
S98T003399	Al	1.07E+02	1.01E+02	5.9
S98T003399	Ca	1.23E+02	1.15E+02	6.7
S98T003399	Fe	1.41E+02	1.29E+02	8.2
S98T003399	Ni	5.98E+01	5.45E+01	8.9
S98T003399	Si	1.52E+01	1.43E+01	6.3
S98T003399	Na	6.12E+02	5.72E+02	6.5
S98T003399	U	1.81E+02	1.70E+02	5.8
S98T003399	Zr	4.92E-02	4.59E-02	6.8

$$\% \text{ Difference} = (\text{ABS}(\text{Sample} - \text{Serial})) / \text{Sample} \times 100$$

Inductively Coupled Plasma/Mass Spectroscopy (ICP/MS)

The liquid subsamples were prepared for analysis by an acid adjustment of the direct subsample. The results for the ICP/MS isotopic analyses are included in Table 1. The solid subsamples were prepared for analysis by performing an acid digest as indicated by an "A" in the A# column in Table 1.

The required analytes were iodine-129 (^{129}I), neptunium-237 (^{237}Np), technetium (^{99}Tc), tin-126 (^{126}Sn), plutonium/ameridium ($\text{Pu}/^{241}\text{Am}$), plutonium isotopes (^{239}Pu , ^{240}Pu and ^{242}Pu) and uranium isotopes (^{233}U , ^{234}U , ^{235}U and ^{238}U). All other analyte results are presented in Appendix A. These results are considered "opportunistic" and do not have customer defined QC parameters. Therefore, any anomalies in those results are not discussed in this report.

The standard recoveries met the requested criteria except for a small percentage of samples analyzed for ^{237}Np (78% - 143.3% recovery), $\text{Pu}/^{241}\text{Am}$ (158.9% recovery) and ^{99}Tc (78.68% recovery). The recovery for ^{99}Tc was only slightly outside of the requested limits of 80% - 120% and was considered acceptable. However, the recoveries for the other two isotopes were outside of the control limits (80% - 120%) requested in the TSAP (Simpson, 1998).

Currently, the ICP/MS is calibrated indirectly for actinides, using a mass response curve derived from the analysis of a 10 ppb thorium-232 (^{232}Th) and ^{238}U standard. This means that the results are semi-quantitative, as the mass response curve only approximates the actual response for any of the specific isotopes. The validity of this approximation is checked using an initial calibration verification (ICV) standard containing ^{232}Th , ^{237}Np , ^{239}Pu , and ^{241}Am at about 10 ppb. Typically, the ^{232}Th recovery is very good (90% - 100%), because, it is in part used to build the calibration curve. For the other three isotopes, recoveries ranging from 50% to 150% are not uncommon.

The actinide method uses an indirect calibration method mainly because of the difficulty in obtaining two separate sets of elemental and isotopic standards that can be used to verify each other. If the customer desires a more accurate result by this method, more development will be required.

High RPDs (>20%) were reported for a small percentage of samples analyzed for ^{235}U and ^{238}U . The RPDs were attributed to sample inhomogeneity and no reanalyses were requested.

The only isotopes for which standard material was available for spike recovery analyses were ^{99}Tc , ^{237}Np , ^{239}Pu and ^{238}U . A small percentage of samples analyzed for these isotopes did not meet the criteria specified in the TSAP for spike recovery (75% - 125%). No reanalyses were requested. As indicated above, recoveries up to 150% are typical for this method.

HNF-1663, REV. 0

The ^{129}I method was new at the time of the analyses and the only standard material available for performing the spike analysis was ^{127}I . Since this parameter (^{127}I) was not set up in the database there are no spike recovery results in Table 1. The calculations for the spike recovery of ^{127}I are contained in the raw data.

Total Inorganic/Organic Carbon Persulfate (TIC/TOC)

TIC/TOC analysis by persulfate oxidation/coulometry was performed in duplicate on direct subsamples. None of the results exceeded the TOC notification limit of 30,000 $\mu\text{gC/g}$ for solids and 30,000 $\mu\text{gC/mL}$ for liquids. The results are included in Table 1.

TIC/TOC was required on the solid subsamples and TOC was required on the liquid subsamples. The results for the liquid TIC are considered "opportunistic", are presented in Appendix A and do not have customer defined QC parameters. Therefore, any anomalies in those results are not discussed in this report.

For all TIC and TOC analyses, an Analysis Report worksheet is included for each sample as raw data. Due to programming limitations with the instrument software, the sample size listed on the worksheet is incorrect. This value was not used in the final calculations and had no bearing on the results in Table 1. The blank is considered a reagent blank. The value was within the acceptance limits and all results were corrected for the concentration found in the blank.

High RPDs (>20%) were reported for five subsamples analyzed for TIC and six subsamples analyzed for TOC. Selected samples were reanalyzed or had triplicate analyses performed with little to no improvement in the RPDs. The high RPDs were attributed to sample inhomogeneity. The raw data for the initial analysis has been included in the data package for informational purposes only and stamped "data not used". The triplicate analysis results are presented in Table 11.

HNF-1663, REV. 0

Table 11. Triplicate Results for TIC/TOC Analyses

Analysis	Labcore ID	Sample Result $\mu\text{gC/g}$ for solids and $\mu\text{gC/mL}$ for liquids	Triplicate Result $\mu\text{gC/g}$ for solids and $\mu\text{gC/mL}$ for liquids	Relative Percent Difference
TIC TOC	S98T002589	5.06E+3 <4.00E+1	3.75E+3 <4.00E+1	29.739% n/a
TIC TOC	S98T002592	5.33E+3 9.90E+2	3.47E+3 1.01E+3	42.273% 2.00%
TIC TOC	S98T002841	4.39E+3 8.24E+3	5.04E+3 1.17E+4	13.786% 34.704%
TIC TOC	S98T002910	1.90E+3 1.69E+3	1.53E+3 <4.00E+1	21.574% n/a
TIC TOC	S98T003011	4.57E+3 2.33E+3	4.49E+3 2.31E+3	1.766% 0.862%
TIC TOC	S98T002737	2.40E+4 3.57E+3	1.32E+4 <4.00E+1	58.065% n/a
TIC TOC	S98T002914	1.36E+4 7.84E+3	1.14E+4 7.77E+3	17.60% 0.897%
TIC TOC	S98T002981	6.82E+3 2.86E+3	9.41E+3 3.51E+2	31.916% 156.275%
TIC TOC	S98T002983	4.09E+3 1.50E+3	3.97E+3 1.76E+3	2.978% 15.951%

HNF-1663, REV. 0

Spike recoveries outside of the 75% to 125% range were reported for five subsamples analyzed for TIC and three subsamples analyzed for TOC. These spike failures were due to sample inhomogeneity and the high concentration of these analytes with respect to the amount of spike standard added. No reanalyses were requested. The standard recoveries were within the required limits of 80%-120%.

Radionuclide Analyses

Total Alpha (AT)

The total alpha (AT) analyses were performed in duplicate on direct subsamples for the liquids as indicated by a blank in the A# column in Table 1. Solid subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion digest is indicated with an "F" in the A# column in Table 1.

All liquid AT results were below the total alpha activity action limit of 61.5 $\mu\text{Ci/mL}$. All solid AT results were below the total alpha activity limit of 28.21 $\mu\text{Ci/g}$ (based on a bulk density of 1.85 g/mL).

High RPDs (>20%) were reported for two of the forty-two subsamples submitted for analysis. The results were near the detection limit which decreased the precision of the analysis. No reanalyses were requested due to the low activity in the samples. More information may be obtained by examining the raw data.

The standard recoveries were within the required limits of 70%-130% and the spike recoveries were within the required limits of 75% to 125%.

Gamma Energy Analysis (GEA)

Gamma energy analysis was performed in duplicate on direct liquid subsamples. The solid subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion digest is indicated with an "F" in the A# column in Table 1.

The QC review discussed in this report is limited to ^{137}Cs . The results for ^{60}Co are considered "opportunistic" and do not have customer defined QC parameters. Therefore, any anomalies in those results are not discussed in this report and the results are presented in Appendix A.

Actual detection limits for GEA analytes are not currently available. The latest GEA software does not report a minimum detectable activity (MDA). If an analyte is reported as "less than", the value reported is the detection limit.

HNF-1663, REV. 0

For the GEA analysis some radionuclides have peak energies that are very close together and cannot always be separated. When this occurs most software will report all isotopes with an energy near the found peak energy. The chemist eliminates the data from the radionuclides that are not present by crossing-out the unused peaks.

High RPDs (>20%) was reported for two of the forty-three subsamples submitted for analysis. These samples were reanalyzed with no improvement in the RPDs. The data for the initial analysis has been included in the data package for informational purposes only and is stamped "data not used". The high RPDs were attributed to sample inhomogeneity and no further reanalysis was requested. The standard recoveries were within the required limits of 80%-120%.

Plutonium 239 (²³⁹Pu)

²³⁹Pu was performed in duplicate on direct liquid subsamples. This analysis was performed inadvertently (it was only required if the total alpha results exceeded the notification limit). However, since these results can be compared with the ICP/MS and total alpha results they were included in Table 1 rather than Appendix A. The standard recoveries were within the required limits of 80% - 120% and the RPDs were less than 20%.

Strontium 90 (⁹⁰Sr)

⁹⁰Sr was performed in duplicate on direct liquid subsamples. The solid subsamples were prepared for analysis by performing a fusion digest in duplicate. The fusion digest is indicated with an "F" in the A# column in Table 1.

There was a small amount of contamination found in the preparation blank. This level of contamination is insignificant when compared to the sample results and does not affect the usability of the data.

High RPDs (>20%) were reported for four of the forty-three subsamples submitted for analysis. Selected samples were reanalyzed. The raw data for the initial analyses have been included in the data package for informational purposes only and stamped as "data not used". The high RPDs were attributed to sample inhomogeneity and no further reanalyses were requested. More information may be obtained by examining the raw data. The standard recoveries were within the required limits of 80% - 120%.

Procedures

Table 12 lists the analytical procedures used for performing the sample analyses. Abbreviations for analyses are defined in the table notes.

Table 12. Analytical Procedures			
Analysis	Sample Portion	Preparation Procedure +	Analysis Procedure
DSC	Solid/Liquid	N/A	LA-514-114 Rev. D-1 LA-514-114 Rev. D-2
TGA	Solid/Liquid	N/A	LA-514-114 Rev. D-1
Bulk Density	Solid	N/A	LO-160-103 Rev. B-0 LA-519-132 Rev. B-0
Sp.G.	Liquid	N/A	LA-510-112 Rev. E-0
NH3	Liquid	N/A	LA-631-001 Rev. C-1 LA-631-001 Rev. D-0
OH	Solid Liquid	LA-504-101 Rev. F-0 N/A	LA-211-102 Rev. D-1
CRVI	Solid Liquid	LA-504-101 Rev. F-0 N/A	LA-265-101 Rev. C-0

Table 12. Analytical Procedures

Analysis	Sample Portion	Preparation Procedure +	Analysis Procedure
IC	Solid Liquid	LA-504-101 Rev. F-0 N/A	LA-533-105 Rev. F-0
ICP	Solid Liquid	LA-505-163 Rev. B-0 LA-549-141 Rev. G-0 N/A	LA-505-161 Rev. C-2 LA-505-161 Rev. C-3
ICP/MS	Solid Liquid	LA-505-163 Rev. B-0 N/A	LA-506-101 Rev. A-0
TIC/TOC	Solid/Liquid	N/A	LA-342-100 Rev. F-1
GEA	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-548-121 Rev. F-0
Pu239	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-953-104 Rev. B-0
Sr90	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-220-101 Rev. E-3
AT	Solid Liquid	LA-549-141 Rev. F-0 N/A	LA-508-101 Rev. G-0

HNF-1663, REV. 0

Abbreviations:

N/A = not applicable (these are direct samples)
DSC = differential scanning calorimetry
TGA = thermogravimetric analysis
OH = hydroxide
NH₃ = ammonia
Sp.G. = specific gravity
IC = ion chromatography
ICP = inductively coupled plasma
TIC/TOC = total inorganic/organic carbon
Sr90 = strontium 90
AT = total alpha
GEA = gamma energy analysis

References

- Simpson, B. C., 1998, *Tank 241-BY-105 Rotary Mode Core Sampling and Analysis Plan*, HNF-2878, Rev. 0, Lockheed Martin Hanford Corporation, Richland, WA 99352.
- Dukelow, G. T., J. W. Hunt, H. Babad, and J. E. Meacham, 1995, *Tank Safety Screening Data Quality Objective*, WHC-SD-WM-SP-004, Rev. 2, Westinghouse Hanford Company, Richland, WA 99352
- Bauer, R. E., and L. P. Jackson, 1998, *Flammable Gas Data Quality Objective*, HNF-SD-WM-DQO-004, Rev. 3, DE&S Hanford, Incorporated, Richland, WA 99352
- Schreiber, R. D., 1997, *Fiscal Year 1997 Memorandum of Understanding for the TWRS Characterization Project*, HNF-SD-WM-SD-023, Rev. 0, Lockheed Martin Hanford Corporation, Richland, WA 99352
- Markel, L. P., 1998, *Quality Assurance Plan for the 222-S Laboratories*, HNF-SD-CP-QAPP-016, Rev. 3, Waste Management of Hanford, Inc., Richland, WA 99352.

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

HNH-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

HNF-1663 REV. 0

BY-105 SAMPLE BREAKDOWN

ATTACHMENT 1

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

BY-105
Core:246
Seg: 1
S98T002475

Attachment 1

BY-105
Core:246
Seg: 2
S98T002476

Attachment 1

HNF-1663 REV. 0

BY-105

Core:246

Seg: 4

S98T002477

Attachment 1

11/11/1993 REV. 0

HNF-1663 REV. 0

BY-105
Core:246
Seg: 8
S98T002625

Attachment 1

BY-105
Core:246
Seg: 9
S98T002473

Attachment 1

BY-105
Core:246
Seg: 9A
S98T002474

Attachment 1

BY-105
Core:246
Seg: FB
S98T002391

Attachment 1

Drainable
Liquid

S98T002391

BSG
TOS
NN3
ALPHA
IC-030N18
IC-030
AC08
IC
TIC10C
OH
CWM
CWM18
PU239
8590
0EN

BY-105
Core:246
Seg: HHF
S98T002392

Attachment 1

Composited:

Segment Grams

1	LH	25.0
2	LH	25.0
4	LH	25.0

Attachment 1

BY-105
Core:246
Composite:Solids

Composited:
Segment: Grams:
9 LH 20.0
9A LH 20.0

Attachment 1

BY-105
Core:246
Composite:Solids

BY-105
Core:246R
Seg: 2R
S98T002626

Attachment 1

HNF-1663 REV. 0

BY-105

Core:246R

Seg: 3R

S98T002627

Attachment 1

HNF-1663 REV. 0

BY-105
Core:246R
Seg: 4R
S98T002628

Attachment 1

BY-105
Core:246R
Seg: 5R
S98T002629

Attachment 1

BY-105

Core:246R

Seg: 6R

S98T002630

Attachment 1

BY-105
Core:246R
Seg: 7R
S98T002631

Attachment 1

BY-105

Core:246R

Seg: 8R

S98T002632

Attachment 1

BY-105
Core:246R
Seg: 9
S98T002633

Attachment 1

BY-105

Core:246R

Composite:

Drainable Liquid

Composited:

Segment: Grams:

2R DL	10.0
3R DL	10.0
4R DL	10.0
5R DL	10.0
6R DL	10.0
7R DL	10.0
8R DL	10.0
9R DL	10.0

BY-105
Core:250
Seg: 1
S98T002634

Attachment 1

|

Sampler
Empty

BY-105
Core:250
Seg: 2
S98T002635

Attachment 1

—

Sampler
Empty

BY-105
Core:250
Seg: 3
S98T002636

Attachment 1

BY-105
Core:250
Seg: 4
S98T002725

Attachment 1

BY-105
Core:250
Seg: 5
S98T002726

Attachment 1

HNF-1663 REV. 0

BY-105
Core:250
Seg: 6
S98T002727

Attachment 1

BY-105
Core:250
Seg: 7
S98T002728

Attachment 1

BY-105
Core:250
Seg: 8
S98T002850

Attachment 1

—

Sampler
Empty

BY-105
Core:250
Seg: 9A
S98T002729

Attachment 1

—

Sampler
Empty

BY-105
Core:250
Seg: 9B
S98T002730

Attachment 1

Upper
Half

S98T002936

Lower
Half

S98T002909

DSC
TGA
TITOC

S98T002914

H2O
DIBST
ICP/MS
ICP/MS
ICP/MS
ICP/MS
ICP/MS

S98T002935

FUSION
ICP
ALPHA
SR90
SEA

S98T002922

ACID
DIBST
ICP/MS

S98T002930

Drainable
Liquid

S98T002937

DSC
TGA
ALPHA
SR90
SEA

S98T002938

ICP
ANIONS
ICP/MS
ICP/MS
ICP/MS
ICP/MS
ICP/MS

S98T002939

ICP/MS
TGA
TITOC
ICP/MS

S98T002940

ARCHIVE

S98T002941

BY-105
Core:250
Composite Solids

Composited:	
Segment:	Grams
3 LH	20.0
4 LH	20.0
5 LH	20.0
6 LH	20.0

Attachment 1

BY-105

Core:251

Seg: 1

S98T002719

Attachment 1

BY-105
Core:251
Seg: 2
S98T002720

Attachment 1

BY-105
Core:251
Seg: 3
S98T002721

Attachment 1

Attachment 1

67

BY-105
Core:251
Seg: 4A
S98T002723

Attachment 1

BY-105

Core:252

Seg: 5

S98T002731

Attachment 1

BY-105
 Core:252
 Seg: 6
 S98T002732

Attachment 1

BY-105
Core:252
Seg: 7
S98T002733

Attachment 1

BY-105
 Core:252
 Seg: 8
 S98T002734

Attachment 1

Attachment 1

BY-105
Core:252
Seg: 9
S98T002735

|

Sampler
Empty

HNF-1663 REV. 0

CORRESPONDENCE

ATTACHMENT 2

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

INTEROFFICE MEMO

LOCKHEED MARTIN

From: Data Assessment and Interpretation
Phone: 373-5915 R2-12
Date: March 2, 1999
Subject: SETTLING RATE STUDY

74B20-99-007

To: F. H. Steen T6-12

cc: R. J. Cash S7-73
D. B. Hardy T6-12
K. L. Powell S3-30
C. M. Seidel S3-30
B. C. Simpson R2-12 *BC*
KMH File/LB

The flammable gas program is in the process of revising the data quality objective to clarify the requirements for the settling rate study. The program has determined that the requirements do not apply to 241-BY-105. Please cancel the settling rate studies for these tank samples.

K. M. Hall,
Manager

dmn

Steen, Franciska H

From: Simpson, Brett C
Sent: Wednesday, December 16, 1998 8:33 AM
To: Steen, Franciska H
Subject: RE: BY-105 Core Composites

Fran--

OK, I know this is late but here goes.

Core 246 Stratum Composites: 246-01 through 246-04 together; 246-05 through 246-07 empty, no problem; 246--08 by itself; 246-09 and 246-09A together.

Core 246R Liquid composite: Equal volumes from each

Core 250: Stratum composites: 250-01 and 250-02 empty, no problem; 250-03 through 250-06 together; 250-07 by itself; 250-08 and 250-09 empty; 250-09A by itself.

Cores 251 and 252, no composites

Hope this isn't too late.

Brett

-----Original Message-----

From: Steen, Franciska H
Sent: Friday, December 04, 1998 8:31 AM
To: Simpson, Brett C
Subject: BY-105 Core Composites

Brett,

Would you please give me direction on compositing the five cores we received for BY-105. Thanks.

Fran

Steen, Franciska H

From: Simpson, Brett C
Sent: Thursday, April 01, 1999 11:23 AM
To: Steen, Franciska H; Hall, Kathleen M
Subject: Typo correction to BY-105 TSAP

Fran--

The "Tank 241-BY-105 Rotary Mode Core Sampling and Analysis Plan" specifies an incorrect laboratory procedure number for measuring bulk density. The correct number is LO-160-103 (instead of LO-160-104). Please proceed with the density analysis using the procedure numbered LO-160-103.

Thanks,

Brett

HNF-1663 REV. 0

SAMPLE DATA SUMMARY

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

Table 1. Data Summary Report.
RY-105 (2)

L: Lower Half of Segment														
R	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicates	Average	RPD %	% Spk	% Det	Limit	Count	Errs
0002583		Bulk Density of Sample	g/mL	n/a	n/a	DRY	n/a	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
0002583		Bulk Density of Sample	g/mL	n/a	n/a	0.00e+00	n/a	0.00e+00	n/a	n/a	n/a	n/a	n/a	n/a
0002589		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	n/a	0.00e+00	n/a	n/a	n/a	n/a	n/a	n/a
0002589		DSC Exotherm on Perkin Elmer	Joules/g	98.56	n/a	0.00e+00	n/a	0.00e+00	n/a	n/a	n/a	n/a	n/a	n/a
0002589		% Water by TGA on Perkin Elmer	%	98.67	n/a	0.00e+00	n/a	0.00e+00	n/a	n/a	n/a	n/a	n/a	n/a
0002589		TIC by AC19/COM/Chemistry	ug/g	99.67	1.400	5.70e+03	3.20e+03	4.13e+03	45.0	101.0	5.00e	2.00	n/a	n/a
0002589		TIC by AC19/COM/Chemistry	ug/g	99.00	10.70	4.50e+01	<4.00e+1	6.48e-01	152	96.80	2.00e+03	3.88E+00	n/a	n/a
0002598		TOC by Persulfate/Oxygen	ug/g	107.7	2.00e-03	1.01e+03	<1.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Strontium-89/90 High Level	uCi/g	99.60	<5.00e-02	1.01e+03	<1.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Aluminum - ICP-Fusion	ug/g	99.80	<5.00e-02	1.01e+03	<1.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Boron - ICP-Fusion	ug/g	100.2	<5.00e-02	1.01e+03	<1.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n/a	n/a	n/a	n/a	n/a
0002598		F Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	n				

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002604	A	Potassium -ICP-Acid Digest	ug/g	97.60	<5.00e-01	<4.61e-02	<4.46e2	n/a	n/a	120.0	460.0	n/a
S98T002604	A	Lithium -ICP-Acid Digest	ug/g	98.00	<1.00e-02	< 9.210	<8.93e0	n/a	n/a	95.00	9.200	n/a
S98T002604	A	Manganese -ICP-Acid Digest	ug/g	94.80	<1.00e-02	< 9.210	<8.93e0	n/a	n/a	95.40	9.200	n/a
S98T002604	A	Sodium -ICP-Acid Digest	ug/g	125.2	8.08e-01	2.65e+05	2.49e+05	2.54e+05	6.71	-2.540e3	92.00	n/a
S98T002604	A	Nickel -ICP-Acid Digest	ug/g	95.40	<2.00e-02	< 18.40	<1.79e1	n/a	n/a	97.80	18.40	n/a
S98T002604	A	Silicon -ICP-Acid Digest	ug/g	33.60	4.25e-01	1.05e+02	91.40	98.20	13.8	98.80	48.10	n/a
S98T002604	A	Uranium -ICP-Acid Digest	ug/g	93.20	<5.00e-01	<4.61e-02	<4.46e2	n/a	n/a	93.60	460.0	n/a
S98T002604	A	Zinc -ICP-Acid Digest	ug/g	89.20	<1.00e-02	< 9.210	<8.93e0	n/a	n/a	97.80	9.200	n/a
S98T002604	A	Zirconium -ICP-Acid Digest	ug/g	93.60	<1.00e-02	< 9.210	<8.93e0	n/a	n/a	94.40	9.200	n/a
S98T002610	W	Formate by IC-Dionex 4000/4500	ug/g	93.75	<6.20e-02	< 70.14	<6.98e1	n/a	n/a	92.37	70.14	n/a
S98T002610	W	Fluoride-IC-Dionex 4000/4500	ug/g	106.9	<1.20e-02	<2.28e+02	<2.27e2	n/a	n/a	104.2	228.5	n/a
S98T002610	W	Chloride-IC-Dionex 4000/4500	ug/g	102.1	<1.70e-02	<3.22e+02	<3.22e2	n/a	n/a	104.2	323.5	n/a
S98T002610	W	Nitrite-IC - Dionex 4000/4500	ug/g	96.51	7.30e-01	3.16e+03	3.23e+03	3.21e+03	2.81	90.95	2.08e+03	n/a
S98T002610	W	Bromide by Ion Chromatograph	ug/g	101.2	<1.25e-01	<2.58e+03	<2.57e3	n/a	n/a	97.30	2.58e+03	n/a
S98T002610	W	Nitrate by IC-Dionex 4000/4500	ug/g	98.14	3.05e-01	6.90e+05	6.49e+05	6.69e+05	6.12	93.30	5.28e+03	n/a
S98T002610	W	Phosphate-IC-Dionex 4000/4500	ug/g	98.90	<1.20e-01	<2.28e+03	<2.27e3	n/a	n/a	94.88	2.28e+03	n/a
S98T002610	W	Sulfate by IC-Dionex 4000/4500	ug/g	100.3	<1.38e-01	<2.63e+03	<2.62e3	n/a	n/a	97.18	2.63e+03	n/a
S98T002610	W	Oxalate-IC-Dionex 4000/4500	ug/g	101.1	<1.05e-01	<2.00e+03	<1.99e3	n/a	n/a	96.86	2.00e+03	n/a

CORE NUMBER: 246
SEGMENT #: 2

Table 1. Data Summary Report.
BY-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec	Det Limit	Count Err
S981002584		Bulk Density of Sample	g/mL	n/a	n/a	DRY	n/a	n/a	0.00	n/a	5.00e-01	n/a
S981002590		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002590		DSC Exotherm Dry Perkin Elmer	Joules/g	98.56	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002590		% Matter by TGA on Perkin Elmer	ug/g	98.67	2.109	2.070	2.070	2.090	1.87	n/a	5.00	n/a
S981002590		TIC by ACD/Countway	ug/g	99.67	1.400	1.50e-04	1.29e-04	1.40e-04	15.1	n/a	40.00	n/a
S981002590		TIC by Perkin Elmer	ug/g	92.00	1.70	58.30	95.90	77.10	48.8	n/a	40.00	n/a
S981002590		Stoic by Perkin Elmer	ug/g	107.7	2.00e-03	2.75e-01	3.40e-03	2.90e-01	10.7	n/a	2.00e-03	2.84e+00
S981002599		F Aluminum - ICP-Fusion	ug/g	99.60	5.00e-02	3.26e+03	3.40e+03	3.33e+03	4.20	n/a	996.0	n/a
S981002599		F Boron - ICP-Fusion	ug/g	99.80	5.00e-02	9.95e+02	9.97e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Barium - ICP-Fusion	ug/g	100.2	5.00e-02	9.95e+02	9.97e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Bismuth - ICP-Fusion	ug/g	99.40	1.00e-01	1.99e+03	1.99e+03	n/a	n/a	n/a	1.99e+03	n/a
S981002599		F Calcium - ICP-Fusion	ug/g	99.00	1.00e-01	1.99e+03	1.99e+03	n/a	n/a	n/a	1.99e+03	n/a
S981002599		F Chromium - ICP-Fusion	ug/g	98.60	1.00e-02	1.99e+02	1.99e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Iron - ICP-Fusion	ug/g	98.20	5.00e-02	9.95e+02	9.97e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Lithium - ICP-Fusion	ug/g	102.0	1.00e-02	1.99e+02	1.99e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Manganese - ICP-Fusion	ug/g	99.00	1.00e-02	1.99e+02	1.99e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Nickel - ICP-Fusion	ug/g	105.0	1.00e-01	2.64e+05	2.65e+05	2.65e+05	0.38	n/a	1.99e+03	n/a
S981002599		F Silicon - ICP-Fusion	ug/g	99.20	1.26e-01	6.53e+02	915.0	785.0	33.1	n/a	996.0	n/a
S981002599		F Strontium - ICP-Fusion	ug/g	98.40	5.00e-02	9.95e+02	9.97e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Uranium - ICP-Fusion	ug/g	95.00	1.00e-02	1.99e+03	1.99e+03	n/a	n/a	n/a	996.0	n/a
S981002599		F Zinc - ICP-Fusion	ug/g	97.80	1.00e-02	1.99e+02	1.99e+02	n/a	n/a	n/a	996.0	n/a
S981002599		F Zirconium - ICP-Fusion	ug/g	105.2	2.60e-03	10.24	10.24	10.27	0.98	n/a	996.0	n/a
S981002599		F Cesium-137 by DEA	ug/g	92.78	2.03e-01	2.45e-01	3.05E-1	10.30	10.27	n/a	3.71e-01	5.00e+02
S981002599		F Alpha of Digest Solid	ug/g	106.6	1.20e-03	4.283	4.42e-01	n/a	n/a	n/a	4.283	n/a
S981002605		A 1129 by ICP/MS Act1819	ug/g	n/a	1.20e-04	4.28e-01	4.42e-1	n/a	n/a	n/a	4.28e-01	n/a
S981002605		A Uranium-235 by ICP/MS Act1819	ug/g	n/a	1.20e-04	4.28e-01	4.42e-1	n/a	n/a	n/a	4.28e-01	n/a
S981002605		A Uranium-238 by ICP/MS Act1819	ug/g	100.7	1.20e-04	4.28e-01	4.42e-1	n/a	n/a	n/a	4.28e-01	n/a
S981002605		A Uranium-236 by ICP/MS Act1819	ug/g	n/a	1.20e-04	4.28e-01	4.42e-1	n/a	n/a	n/a	4.28e-01	n/a
S981002605		A Uranium-234 by ICP/MS Act1819	ug/g	99.00	1.20e-04	4.28e-01	4.42e-1	n/a	n/a	n/a	4.28e-01	n/a
S981002605		A Uranium-233 by ICP/MS Act1819	ug/g	91.71	2.20e-03	3.18e-01	3.35e-01	3.26e-01	5.21	n/a	7.10e-02	n/a
S981002605		A Neptunium-237 by ICP/MS Act1819	ug/g	115.9	3.45e-03	1.12e-01	1.10e-1	1.10e-1	n/a	n/a	1.2e-01	n/a
S981002605		A Plutonium-239 by ICP/MS Act1819	ug/g	101.0	3.45e-03	1.12e-01	1.10e-1	1.10e-1	n/a	n/a	1.52e-01	n/a
S981002605		A Plutonium-240 by ICP/MS Act1819	ug/g	n/a	3.45e-03	1.12e-01	1.10e-1	1.10e-1	n/a	n/a	1.18e-01	n/a
S981002605		A Plutonium-241 by ICP/MS Act1819	ug/g	n/a	3.45e-03	1.12e-01	1.10e-1	1.10e-1	n/a	n/a	1.18e-01	n/a
S981002605		A Plutonium-242 by ICP/MS Act1819	ug/g	103.9	3.45e-03	1.12e-01	1.10e-1	1.10e-1	n/a	n/a	1.18e-01	n/a
S981002605		A Americium-243 by ICP/MS Act1819	ug/g	97.40	9.70e-02	3.36e+03	3.04e+03	3.20e+03	10.0	n/a	44.40	n/a
S981002605		A Americium - ICP-Acid Digest	ug/g	106.6	4.88e-01	1.26e+02	118.0	122.0	6.56	n/a	44.40	n/a
S981002605		A Boron - ICP-Acid Digest	ug/g	98.80	5.00e-02	9.95e+02	9.97e+02	n/a	n/a	n/a	44.40	n/a
S981002605		A Barium - ICP-Acid Digest	ug/g	94.40	1.00e-01	1.99e+03	1.99e+03	n/a	n/a	n/a	89.00	n/a
S981002605		A Bismuth - ICP-Acid Digest	ug/g	106.4	1.77e-01	88.80	91.0	88.80	173.0	n/a	89.00	n/a
S981002605		A Calcium - ICP-Acid Digest	ug/g	93.60	1.00e-02	99.10	112.0	105.5	12.2	n/a	8.900	n/a
S981002605		A Chromium - ICP-Acid Digest	ug/g	93.20	3.30e-02	44.40	44.58e1	n/a	n/a	n/a	44.40	n/a
S981002605		A Iron - ICP-Acid Digest	ug/g	93.20	3.30e-02	44.40	44.58e1	n/a	n/a	n/a	44.40	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002605	A	Potassium -ICP-Acid Digest	ug/g	97.60	<5.00e-01	<4.44e-02	508.0	n/a	n/a	n/a	444.0	n/a
S98T002605	A	Lithium -ICP-Acid Digest	ug/g	98.00	<1.00e-02	< 8.880	<9.16e0	n/a	n/a	n/a	8.900	n/a
S98T002605	A	Manganese -ICP-Acid Digest	ug/g	94.80	<1.00e-02	< 8.880	<9.16e0	n/a	n/a	n/a	8.900	n/a
S98T002605	A	Sodium -ICP-Acid Digest	ug/g	125.2	8.08e-01	2.51e+05	2.60e+05	2.56e+05	3.52	n/a	89.00	n/a
S98T002605	A	Nickel -ICP-Acid Digest	ug/g	95.40	<2.00e-02	< 17.80	<1.83e1	n/a	n/a	n/a	17.80	n/a
S98T002605	A	Silicon -ICP-Acid Digest	ug/g	33.60	4.25e-01	96.40	82.10	89.25	16.0	n/a	44.40	n/a
S98T002605	A	Uranium -ICP-Acid Digest	ug/g	89.20	<5.00e-01	<4.44e-02	<4.58e2	n/a	n/a	n/a	444.0	n/a
S98T002605	A	Zinc -ICP-Acid Digest	ug/g	89.20	<1.00e-02	< 8.880	<9.16e0	n/a	n/a	n/a	8.900	n/a
S98T002605	A	Zirconium -ICP-Acid Digest	ug/g	93.75	<6.20e-02	< 1.14e+02	<4.36e0	n/a	n/a	n/a	8.900	n/a
S98T002611	W	Formate by IC-Dionex 4000/4500	ug/g	106.9	<1.20e-02	4.27e+02	436.0	431.4	5.96	n/a	232.3	n/a
S98T002611	W	Fluoride-IC-Dionex 4000/4500	ug/g	102.1	<1.70e-02	<3.29e+02	<3.30e2	n/a	n/a	n/a	329.1	n/a
S98T002611	W	Nitrite-IC - Dionex 4000/4500	ug/g	96.51	7.30e-01	5.20e+03	5.70e+03	5.45e+03	9.17	n/a	2.09e+03	n/a
S98T002611	W	Bromide by Ion Chromatograph	ug/g	101.2	<1.25e-01	<2.42e+03	<2.42e3	n/a	n/a	n/a	2.42e+03	n/a
S98T002611	W	Nitrate by IC-Dionex 4000/4500	ug/g	98.14	3.05e-01	6.16e+03	5.75e+05	5.95e+05	7.23	n/a	2.69e+03	n/a
S98T002611	W	Phosphate-IC-Dionex 4000/4500	ug/g	98.90	<1.20e-01	<2.32e+03	<2.35e3	n/a	n/a	n/a	2.32e+03	n/a
S98T002611	W	Sulfate by IC-Dionex 4000/4500	ug/g	100.3	<1.38e-01	<2.67e+03	<2.68e3	n/a	n/a	n/a	2.67e+03	n/a
S98T002611	W	Oxalate-IC-Dionex 4000/4500	ug/g	101.1	<1.05e-01	<2.03e+03	<2.04e3	n/a	n/a	n/a	2.03e+03	n/a

Table 1. Data Summary Report.
BT-105 (2)CORE NUMBER: 246
SEGMENT #: 3

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Error
S98T002591		DSC Exotherm Dry	Joules/g Dry	n/a	n/a	5.50	6.280	6.905	18.1	n/a	n/a
S98T002591		DSC Exotherm on Perkin Elmer	Joules/g	98.84	n/a	6.70	5.630	5.950	10.8	n/a	n/a
S98T002591		% Water by TGA on Perkin Elmer	%	99.29	n/a	18.40	14.77	16.69	23.0	n/a	n/a
S98T002591		TIC by Acid/Coulometry	ug/g	99.67	1.400	3.99e+04	4.39e+04	4.18e+04	9.80	5.000	n/a
S98T002591		TIC by Persulfate/Coulometry	ug/g	92.00	10.70	3.76e+03	3.90e+03	3.83e+03	3.66	40.00	n/a
S98T002591		TIC by Persulfate/Coulometry	ug/g	107.7	2.00e-03	9.750	10.90	10.32	11.1	9.00e-01	9.54E-01
S98T002600	F	Strontium-89/90 High Level	uc/g	99.60	<5.00e-02	1.63e-02	2.43e+04	2.03e+04	39.4	n/a	n/a
S98T002600	F	Aluminum - ICP-Fusion	ug/g	99.80	<5.00e-02	1.02e-02	<1.02e-02	n/a	n/a	1.02e+03	n/a
S98T002600	F	Boron - ICP-Fusion	ug/g	100.2	<5.00e-02	1.02e-02	<1.02e-02	n/a	n/a	1.02e+03	n/a
S98T002600	F	Barium - ICP-Fusion	ug/g	99.40	<1.00e-01	2.03e-02	<2.03e-02	n/a	n/a	2.03e+03	n/a
S98T002600	F	Bismuth - ICP-Fusion	ug/g	99.00	<1.00e-01	2.03e-02	<2.03e-02	n/a	n/a	2.03e+03	n/a
S98T002600	F	Calcium - ICP-Fusion	ug/g	98.60	<1.00e-02	4.78e-02	4.57	4.67	4.69	203.0	n/a
S98T002600	F	Chromium - ICP-Fusion	ug/g	98.20	<5.00e-02	1.02e-02	<1.02e-02	n/a	n/a	1.02e+03	n/a
S98T002600	F	Iron - ICP-Fusion	ug/g	102.0	<1.00e-02	2.03e-02	<2.03e-02	n/a	n/a	203.0	n/a
S98T002600	F	Lithium - ICP-Fusion	ug/g	99.00	<1.00e-01	2.03e-02	<2.03e-02	n/a	n/a	2.03e+03	n/a
S98T002600	F	Manganese - ICP-Fusion	ug/g	105.0	<1.00e-01	2.03e-02	2.70e+05	2.75e+05	3.28	n/a	n/a
S98T002600	F	Sodium - ICP-Fusion	ug/g	99.20	<1.00e-01	2.03e-02	6.61	584.0	19.5	1.02e+03	n/a
S98T002600	F	Nickel - ICP-Fusion	ug/g	98.40	<5.00e-02	1.02e-02	4.01e+03	3.06e+03	62.5	n/a	n/a
S98T002600	F	Silicon - ICP-Fusion	ug/g	98.60	<5.00e-02	1.02e-02	<1.02e-02	n/a	n/a	203.0	n/a
S98T002600	F	Zinc - ICP-Fusion	ug/g	95.00	<1.00e-02	2.03e-02	<2.03e-02	n/a	n/a	1.02e+04	n/a
S98T002600	F	Zinc - ICP-Fusion	ug/g	97.80	<1.00e-02	2.03e-02	<2.03e-02	n/a	n/a	203.0	n/a
S98T002600	F	Zirconium - ICP-Fusion	ug/g	105.2	<7.60e-03	2.03e-02	21.80	21.27	5.18	n/a	0.340
S98T002600	F	Cesium-137 by GEA	uc/g	92.78	<7.60e-03	2.03e-02	<1.59e-1	n/a	n/a	3.79e-01	5.00E+02
S98T002600	F	Alpha of Digested Solid	uc/g	106.6	<1.20e-03	4.52e-02	<4.52e-02	n/a	n/a	4.52e-01	n/a
S98T002600	A	1129 by ICP/MS Acid/DigLA505159	ug/g	n/a	<1.20e-04	4.52e-01	<4.52e-01	n/a	n/a	4.52e-01	n/a
S98T002600	A	Uranium-233 by ICP/MS Acid/DigLA505159	ug/g	n/a	<1.20e-04	4.52e-01	<4.52e-01	n/a	n/a	4.52e-01	n/a
S98T002600	A	Uranium-234 by ICP/MS Acid/DigLA505159	ug/g	100.7	<1.20e-04	4.52e-01	2.070	1.942	13.4	n/a	n/a
S98T002600	A	Uranium-235 by ICP/MS Acid/DigLA505159	ug/g	n/a	<1.20e-04	4.52e-01	<4.52e-01	n/a	n/a	6.03e-01	n/a
S98T002600	A	Uranium-238 by ICP/MS Acid/DigLA505159	ug/g	99.00	<1.20e-04	4.52e-01	<4.52e-01	n/a	n/a	4.52e-01	n/a
S98T002600	A	Uranium-238 by ICP/MS Acid/DigLA505159	ug/g	91.71	<1.20e-03	4.52e-01	6.150	5.81e-01	12.0	n/a	7.50e-02
S98T002600	A	IC99 by ICP/MS Acid/DigLA505159	ug/g	n/a	<1.20e-03	4.52e-01	<4.52e-01	n/a	n/a	5.652	n/a
S98T002600	A	Sn126 by ICP/MS Acid/DigLA505159	ug/g	115.9	<4.52e-03	4.52e-01	<4.52e-01	n/a	n/a	1.18e-01	n/a
S98T002600	A	Neptunium-237 by ICP/MS	ug/g	101.0	<3.00e-03	4.52e-01	<4.52e-01	n/a	n/a	1.06e-01	n/a
S98T002600	A	Plutonium-239 by ICP/MS	ug/g	n/a	<4.70e-03	4.52e-01	<4.52e-01	n/a	n/a	1.23e-01	n/a
S98T002600	A	Plutonium-240 by ICP/MS	ug/g	n/a	<4.70e-03	4.52e-01	<4.52e-01	n/a	n/a	1.23e-01	n/a
S98T002600	A	Plutonium-242 by ICP/MS	ug/g	103.9	<3.59e-03	4.52e-01	<4.52e-01	n/a	n/a	1.23e-01	n/a
S98T002600	A	Pu/Am-241 by ICP/MS	ug/g	97.60	9.70e-02	1.74e+04	1.89e+04	1.79e+04	5.59	n/a	46.80
S98T002600	A	Aluminum - ICP-Acid Digest	ug/g	105.6	4.88e-01	1.37e+02	124.0	130.5	9.96	n/a	46.80
S98T002600	A	Boron - ICP-Acid Digest	ug/g	96.80	<5.00e-02	4.52e-01	<4.52e-01	n/a	n/a	46.80	n/a
S98T002600	A	Barium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	2.03e-02	<2.03e-02	n/a	n/a	93.50	n/a
S98T002600	A	Bismuth - ICP-Acid Digest	ug/g	106.4	1.77e-01	2.07e+02	221.0	214.0	6.54	n/a	93.50
S98T002600	A	Calcium - ICP-Acid Digest	ug/g	95.60	<1.00e-02	4.52e-02	52.0	488.5	15.8	n/a	9.350
S98T002600	A	Chromium - ICP-Acid Digest	ug/g	95.20	<5.00e-02	4.52e-02	178.0	675.0	12.7	n/a	46.80
S98T002600	A	Iron - ICP-Acid Digest	ug/g	97.60	<5.00e-01	4.76e+02	677.0	576.5	34.9	n/a	468.0

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec. %	Det Limit	Count Err%
S981002605	A	Lithium - ICP-Acid Digest	ug/g	98.00	<1.00e-02	18.00	20.30	19.15	12.0	n/a	9.350	n/a
S981002605	A	Manganese - ICP-Acid Digest	ug/g	94.80	<1.00e-02	26.00	28.10	27.05	7.76	n/a	9.350	n/a
S981002605	A	Sodium - ICP-Acid Digest	ug/g	125.2	8.08e-01	2.72e+05	2.75e+05	2.74e+05	1.10	n/a	93.50	n/a
S981002605	A	Nickel - ICP-Acid Digest	ug/g	95.40	<2.00e-02	87.00	95.70	91.35	9.52	n/a	18.80	n/a
S981002605	A	Silicon - ICP-Acid Digest	ug/g	33.60	4.25e-01	3.41e+02	301.0	321.0	12.5	n/a	46.80	n/a
S981002605	A	Uranium - ICP-Acid Digest	ug/g	95.20	<5.00e-01	44.69e+02	44.62e2	n/a	n/a	n/a	468.0	n/a
S981002605	A	Zinc - ICP-Acid Digest	ug/g	89.20	<1.00e-02	9.370	9.270	n/a	n/a	n/a	9.350	n/a
S981002612	A	Zinc - ICP-Acid Digest	ug/g	93.60	<1.00e-02	9.370	49.23e0	n/a	n/a	n/a	9.350	n/a
S981002612	A	Formate by IC-Dionex 4000/4500	ug/g	98.32	6.20e-02	2.11e+02	263.0	236.8	21.9	n/a	70.98	n/a
S981002612	A	Fluoride-IC-Dionex 4000/4500	ug/g	106.9	<1.20e-02	7.27e+03	8.58e+03	7.92e+03	16.5	n/a	231.3	n/a
S981002612	A	Chloride-IC-Dionex 4000/4500	ug/g	102.1	<1.70e-02	33.28e+02	339.0	n/a	n/a	n/a	327.6	n/a
S981002612	A	Nitrite-IC - Dionex 4000/4500	ug/g	101.2	5.04e-01	4.89e+03	7.11e+03	6.00e+03	37.0	n/a	845.0	n/a
S981002612	A	Bromide by Ion Chromatograph	ug/g	99.83	<1.25e-01	9.78e+02	9.80e2	n/a	n/a	n/a	977.9	n/a
S981002612	A	Nitrate by IC-Dionex 4000/4500	ug/g	99.32	<1.50e-01	3.70e+05	1.88e+05	2.44e+05	45.9	n/a	1.09e+03	n/a
S981002612	A	Phosphate-IC-Dionex 4000/4500	ug/g	98.90	<1.20e-01	1.76e+04	1.63e+04	1.69e+04	6.53	n/a	2.31e+03	n/a
S981002612	A	Sulfate by IC-Dionex 4000/4500	ug/g	101.6	<1.38e-01	1.17e+04	1.54e+04	1.36e+04	27.3	n/a	1.08e+03	n/a
S981002612	A	Oxalate-IC-Dionex 4000/4500	ug/g	101.1	<1.05e-01	1.23e+04	1.41e+04	1.32e+04	13.6	n/a	2.02e+03	n/a

Table 1. Data Summary Report.
81-105 (2)

CORE NUMBER: 246
SEGMENT #: 4

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002586		Bulk Density of sample	g/mL	n/a	n/a	DRY	n/a	n/a	n/a	5.00e-01	n/a
S98T002592		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	1.630	1.620	1.728	11.0	n/a	n/a
S98T002592		DSC Exotherm on Perkin Elmer	Joules/g	98.84	n/a	2.390	2.380	1.492	11.3	n/a	n/a
S98T002592		% Water by TGA on Perkin Elmer	%	99.29	n/a	2.631	2.620	2.372	21.8	n/a	n/a
S98T002592		TIC by Acid/Coulometry	ug/g	98.50	2.700	5.35e+03	5.80e+03	5.58e+03	8.79	5.000	n/a
S98T002592		TIC by Persulfate/Coulometry	ug/g	92.67	2.700	1.70	1.670	878.5	25.4	40.00	1.19e+00
S98T002601		F Strontium-89/90 High Level	uCi/g	101.9	5.40e-02	1.500	1.470	1.495	2.02	7.77e-04	1.19e+00
S98T002601		F Aluminum - ICP-Fusion	ug/g	100.0	<5.00e-02	1.82e+04	1.71e+04	1.76e+04	6.25	992.0	n/a
S98T002601		F Boron - ICP-Fusion	ug/g	100.4	<5.00e-02	4.95e+02	4.95e+02	n/a	n/a	992.0	n/a
S98T002601		F Barium - ICP-Fusion	ug/g	101.2	<5.00e-02	4.95e+02	4.95e+02	n/a	n/a	992.0	n/a
S98T002601		F Bismuth - ICP-Fusion	ug/g	101.2	<1.00e-01	4.19e+03	4.19e+03	n/a	n/a	1.98e+03	n/a
S98T002601		F Calcium - ICP-Fusion	ug/g	99.80	<1.00e-01	4.19e+03	4.19e+03	n/a	n/a	1.98e+03	n/a
S98T002601		F Chromium - ICP-Fusion	ug/g	99.80	<1.00e-01	4.19e+03	4.19e+03	n/a	n/a	992.0	n/a
S98T002601		F Iron - ICP-Fusion	ug/g	98.80	<5.00e-02	4.95e+02	4.95e+02	n/a	n/a	992.0	n/a
S98T002601		F Lithium - ICP-Fusion	ug/g	103.0	<1.00e-02	1.99e+02	1.99e+02	n/a	n/a	198.0	n/a
S98T002601		F Manganese - ICP-Fusion	ug/g	99.20	<1.00e-02	1.99e+02	2.45e+05	2.43e+05	1.65	1.98e+03	n/a
S98T002601		F Nickel - ICP-Fusion	ug/g	98.60	<5.00e-02	1.55e+02	1.31e+03	1.33e+03	16.3	992.0	n/a
S98T002601		F Silicon - ICP-Fusion	ug/g	99.20	<5.00e-02	4.95e+02	4.95e+02	n/a	n/a	992.0	n/a
S98T002601		F Uranium - ICP-Fusion	ug/g	99.20	<5.00e-02	4.95e+02	4.95e+02	n/a	n/a	9.92e+03	n/a
S98T002601		F Zinc - ICP-Fusion	ug/g	99.40	<1.00e-02	1.99e+02	1.99e+02	n/a	n/a	198.0	n/a
S98T002601		F Zirconium - ICP-Fusion	ug/g	98.40	<1.00e-02	1.99e+02	1.99e+02	n/a	n/a	198.0	n/a
S98T002601		F Cesium-137 by GEA	uCi/g	100.2	<2.28e-01	8.769	8.830	8.800	0.68	n/a	1.00
S98T002601		F Alpha of Digested Solid	uCi/g	100.2	<2.28e-01	2.28e-01	2.28e-01	n/a	n/a	5.31e-01	5.00e+02
S98T002607		A U-235 by ICP/MS Acid/LA305159	ug/g	109.0	<1.00e-03	4.483	4.482e-01	n/a	n/a	4.683	n/a
S98T002607		A Uranium-235 by ICP/MS Acid/LA305159	ug/g	n/a	<1.00e-03	4.68e-01	4.68e-01	n/a	n/a	4.68e-01	n/a
S98T002607		A Uranium-234 by ICP/MS Acid/LA305159	ug/g	n/a	<1.00e-03	4.68e-01	4.68e-01	n/a	n/a	4.68e-01	n/a
S98T002607		A Uranium-235 by ICP/MS Acid/LA305159	ug/g	110.6	<1.00e-03	5.31e-01	5.34e-01	5.42e-01	4.24	6.68e-01	n/a
S98T002607		A Uranium-236 by ICP/MS Acid/LA305159	ug/g	n/a	<1.00e-03	4.68e-01	4.68e-01	n/a	n/a	6.25e-01	n/a
S98T002607		A Uranium-238 by ICP/MS Acid/LA305159	ug/g	99.6	<1.00e-03	4.68e-01	4.68e-01	n/a	n/a	6.25e-01	n/a
S98T002607		A Tc-99 by ICP/MS Acid/LA305159	ug/g	93.64	<1.00e-03	3.20e-01	2.75e-01	2.96e-01	15.9	7.80e-02	n/a
S98T002607		A Sr-87 by ICP/MS Acid/LA305159	ug/g	100.2	<1.00e-03	4.68e-01	4.68e-01	n/a	n/a	5.855	n/a
S98T002607		A Neptunium-237 by ICP/MS	ug/g	100.2	<1.00e-03	1.82e-01	1.82e-01	n/a	n/a	1.82e-01	n/a
S98T002607		A Plutonium-239 by ICP/MS	ug/g	104.0	<1.00e-03	1.35e-01	1.35e-01	n/a	n/a	1.35e-01	n/a
S98T002607		A Plutonium-240 by ICP/MS	ug/g	n/a	<1.00e-03	1.35e-01	1.35e-01	n/a	n/a	1.35e-01	n/a
S98T002607		A Plutonium-242 by ICP/MS	ug/g	110.4	<1.00e-03	1.35e-01	1.35e-01	n/a	n/a	1.35e-01	n/a
S98T002607		A Pu/Am-241 by ICP/MS	ug/g	100.4	<1.00e-03	1.35e-01	1.35e-01	n/a	n/a	1.35e-01	n/a
S98T002607		A Aluminum - ICP-Acid Digest	ug/g	93.40	<1.00e-02	1.45e+02	1.78e+04	1.92e+04	14.6	38.80	n/a
S98T002607		A Boron - ICP-Acid Digest	ug/g	100.2	<5.00e-02	4.95e+02	4.95e+02	111.3	60.2	38.80	n/a
S98T002607		A Barium - ICP-Acid Digest	ug/g	88.20	<1.00e-02	4.95e+02	4.95e+02	n/a	n/a	38.80	n/a
S98T002607		A Bismuth - ICP-Acid Digest	ug/g	99.40	<1.00e-02	4.95e+02	4.95e+02	n/a	n/a	7.60	n/a
S98T002607		A Calcium - ICP-Acid Digest	ug/g	89.40	<1.00e-02	1.78e+02	1.78e+02	169.0	10.7	7.60	n/a
S98T002607		A Chromium - ICP-Acid Digest	ug/g	87.40	<5.00e-02	69.20	61.60	65.40	11.6	38.80	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err-X
S98T002607	A	Potassium -ICP-Acid Digest	ug/g	91.80	<5.00e-01	<3.88e+02	<4.00e2	n/a	n/a	n/a	388.0	n/a
S98T002607	A	Lithium -ICP-Acid Digest	ug/g	99.80	<1.00e-02	< 7.70	<7.99e0	n/a	n/a	n/a	7.760	n/a
S98T002607	A	Manganese -ICP-Acid Digest	ug/g	87.80	<1.00e-02	< 7.70	<7.99e0	n/a	n/a	n/a	7.760	n/a
S98T002607	A	Sodium -ICP-Acid Digest	ug/g	121.6	9.39e-01	2.54e+05	2.42e+05	2.48e+05	4.84	n/a	77.60	n/a
S98T002607	A	Nickel -ICP-Acid Digest	ug/g	87.40	<2.00e-02	19.00	19.80	19.40	4.12	n/a	15.50	n/a
S98T002607	A	Silicon -ICP-Acid Digest	ug/g	157.4	1.100	2.76e+02	163.0	219.5	51.5	n/a	38.80	n/a
S98T002607	A	Uranium -ICP-Acid Digest	ug/g	95.40	<5.00e-01	<3.88e+02	<4.00e2	n/a	n/a	n/a	388.0	n/a
S98T002607	A	Zinc -ICP-Acid Digest	ug/g	76.80	<1.00e-02	< 7.70	<7.99e0	n/a	n/a	n/a	7.760	n/a
S98T002607	A	Zirconium -ICP-Acid Digest	ug/g	91.80	<1.00e-02	< 7.70	<7.99e0	n/a	n/a	n/a	7.760	n/a
S98T002613	M	Formate by IC-Dioxex 4000/4500	ug/g	106.7	<6.20e-02	<2.63e+02	<2.58e2	n/a	n/a	n/a	263.3	n/a
S98T002613	M	Fluoride-IC-Dioxex 4000/4500	ug/g	108.1	<1.20e-02	2.98e+02	558.0	428.1	60.7	n/a	99.51	n/a
S98T002613	M	Chloride-IC-Dioxex 4000/4500	ug/g	105.8	<1.70e-02	3.03e+03	240.0	248.8	7.23	n/a	141.0	n/a
S98T002613	M	Nitrite-IC - Dioxex 4000/4500	ug/g	90.65	<1.08e-01	3.03e+03	2.95e+03	2.99e+03	2.68	n/a	895.4	n/a
S98T002613	M	Bromide by Ion Chromatograph	ug/g	99.15	<1.25e-01	<1.04e+03	<1.02e3	n/a	n/a	n/a	1.04e+03	n/a
S98T002613	M	Nitrate by IC-Dioxex 4000/4500	ug/g	93.49	<1.39e-01	7.13e+05	6.47e+05	6.80e+05	9.71	n/a	2.84e+03	n/a
S98T002613	M	Phosphate-IC-Dioxex 4000/4500	ug/g	104.2	<1.39e-01	2.53e+03	2.57e+03	2.55e+03	1.57	n/a	995.1	n/a
S98T002613	M	Sulfate by IC-Dioxex 4000/4500	ug/g	100.8	<1.38e-01	1.80e+03	1.50e+03	1.65e+03	18.2	n/a	1.14e+03	n/a
S98T002613	M	Oxalate-IC-Dioxex 4000/4500	ug/g	109.4	<1.05e-01	2.94e+03	2.62e+03	2.78e+03	11.5	n/a	870.4	n/a

Table 1. Data Summary Report.
8Y-105 (2)CORE NUMBER: 246
SEGMENT #: 8

SEGMENT PORTION: L Lower Half of Segment

PORTION: L Lower Half of Segment													
Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%	
S981002710		Bulk Density of Sample	g/mL	n/a	n/a	1.740	n/a	n/a	n/a	n/a	5.00e-01	n/a	
S981002710		DSC Exotherm DTA Calculated	Joules/g Dry	n/a	n/a	29.57	25.24	27.41	15.8	n/a	n/a	n/a	
S981002711		DSC Exotherm on Perkin Elmer	Joules/g	100.6	n/a	20.76	15.90	19.24	40.9	n/a	n/a	n/a	
S981002711		% Water by TGA on Perkin Elmer	%	98.99	n/a	23.70	35.90	29.80	24.5	0.00e+00	5.000	n/a	
S981002711		TIC by AGL/Conductivity	ug/g	99.50	1.100	3.12e+04	2.44e+04	2.78e+04	8.01	202.0	40.00	6.87E-01	
S981002711		TOC by Persulfate/Coulometry	ug/g	91.67	7.200	6.25e+03	5.75e+03	5.99e+03	8.01	202.0	40.00	6.87E-01	
S981002711		F Strong Ion - ICP-Fusion	ug/g	97.11	6.70e-02	46.40	1.64e+04	1.82e+04	20.3	n/a	1.04e+03	n/a	
S981002712		F Aluminum - ICP-Fusion	ug/g	99.20	5.00e-02	2.01e+03	<1.05e3	n/a	n/a	n/a	1.04e+03	n/a	
S981002712		F Boron - ICP-Fusion	ug/g	104.6	5.00e-02	<1.04e+03	<1.05e3	n/a	n/a	n/a	2.08e+03	n/a	
S981002712		F Barium - ICP-Fusion	ug/g	99.00	<1.00e-01	<2.08e+03	<2.09e3	n/a	n/a	n/a	2.08e+03	n/a	
S981002712		F Bismuth - ICP-Fusion	ug/g	101.4	<1.00e-01	1.27e+04	1.21e+04	1.24e+04	8.84	n/a	208.0	n/a	
S981002712		F Calcium - ICP-Fusion	ug/g	104.0	<1.00e-02	2.94e+03	2.37e+03	2.66e+03	22.1	n/a	1.04e+03	n/a	
S981002712		F Chromium - ICP-Fusion	ug/g	101.2	<1.00e-02	6.88e+03	5.94e+03	6.41e+03	14.7	n/a	208.0	n/a	
S981002712		F Iron - ICP-Fusion	ug/g	97.80	<1.00e-02	<2.08e+02	<2.09e2	n/a	n/a	n/a	208.0	n/a	
S981002712		F Lithium - ICP-Fusion	ug/g	103.2	<1.00e-02	<2.08e+02	<2.09e2	n/a	n/a	n/a	2.08e+03	n/a	
S981002712		F Manganese - ICP-Fusion	ug/g	100.0	<1.00e-01	2.45e+05	2.13e+05	2.29e+05	14.0	n/a	418.0	n/a	
S981002712		F Sodium - ICP-Fusion	ug/g	102.6	<1.00e-01	8.19e+03	1.36e+04	1.09e+04	49.8	n/a	1.04e+03	n/a	
S981002712		F Nickel - ICP-Fusion	ug/g	104.2	<1.00e-02	8.94e+03	6.81e+03	7.88e+03	27.0	n/a	1.04e+03	n/a	
S981002712		F Silicon - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.08e+02	<2.09e2	n/a	n/a	n/a	208.0	n/a	
S981002712		F Uranium - ICP-Fusion	ug/g	99.40	<1.00e-02	<2.08e+02	<2.09e2	n/a	n/a	n/a	208.0	n/a	
S981002712		F Zinc - ICP-Fusion	ug/g	103.4	<1.00e-02	<2.08e+02	<2.09e2	n/a	n/a	n/a	208.0	n/a	
S981002712		F Zirconium - ICP-Fusion	ug/g	89.42	1.87e-01	1.12e+02	92.80	102.3	18.8	n/a	3.42e-01	5.00E-02	
S981002712		F Deslign 137 by GEA	ug/g	109.0	<2.45e-01	<1.72e-01	<1.56e-1	n/a	n/a	n/a	4.786	n/a	
S981002712		F 1125 by ICP/MS Acid/Dig LA-505159	ug/g	109.0	<1.20e-03	<4.78e-01	<4.69e-1	n/a	n/a	n/a	4.79e-01	n/a	
S981002712		F Uranium-233 by ICP/MS Acid/Dig159	ug/g	n/a	<1.32e-04	<4.79e-01	<4.69e-1	n/a	n/a	n/a	4.79e-01	n/a	
S981002713		A Uranium-234 by ICP/MS Acid/Dig159	ug/g	98.59	<1.32e-04	24.67	30.80	27.73	22.0	n/a	6.38e-01	n/a	
S981002713		A Uranium-235 by ICP/MS Acid/Dig159	ug/g	94.50	<1.32e-04	6.74e-01	8.41e-01	7.59e-01	22.0	n/a	6.38e-01	n/a	
S981002713		A Uranium-238 by ICP/MS Acid/Dig159	ug/g	93.66	<1.32e-04	3.58e+03	4.47e+03	3.99e+03	20.8	n/a	8.00e-02	n/a	
S981002713		A Uranium-238 by ICP/MS Acid/Dig LA-505159	ug/g	93.66	4.91e-04	<1.452	1.980	1.516	8.58	n/a	8.00e-02	n/a	
S981002713		A 1CP9 by ICP/MS Acid/Dig LA-505159	ug/g	118.9	<2.31e-05	<6.97e-01	<1.64e-1	n/a	n/a	n/a	1.67e-01	n/a	
S981002713		A Neptunium-237 by ICP/MS	ug/g	99.60	<4.01e-05	<2.80e-01	<2.85e-1	n/a	n/a	n/a	2.90e-01	n/a	
S981002713		A Plutonium-239 by ICP/MS	ug/g	n/a	<1.12e-05	<8.09e-02	<7.95e-2	n/a	n/a	n/a	8.10e-02	n/a	
S981002713		A Plutonium-240 by ICP/MS	ug/g	103.9	<3.45e-05	<2.50e-01	<2.45e-1	n/a	n/a	n/a	2.50e-01	n/a	
S981002713		A Plutonium-242 by ICP/MS	ug/g	96.40	<1.85e-05	<1.34e-01	<1.31e-1	n/a	n/a	n/a	1.34e-01	n/a	
S981002713		A Pu/Ac-244 by ICP/MS	ug/g	96.40	1.42e-05	2.02e+04	2.11e+04	2.06e+04	4.36	n/a	49.60	n/a	
S981002713		A Aluminum - ICP-Acid Digest	ug/g	107.4	7.27e-01	1.38e+02	50.00	58.20	34.10	15.2	n/a	49.60	n/a
S981002713		A Boron - ICP-Acid Digest	ug/g	93.00	<1.00e-02	2.81e+02	328.0	304.5	15.4	n/a	99.00	n/a	
S981002713		A Barium - ICP-Acid Digest	ug/g	91.80	<1.57e-01	1.39e+04	1.72e+04	1.56e+04	2.12	n/a	9.900	n/a	
S981002713		A Bismuth - ICP-Acid Digest	ug/g	105.0	<1.00e-02	2.95e+03	3.06e+03	3.06e+03	3.66	n/a	99.00	n/a	
S981002713		A Calcium - ICP-Acid Digest	ug/g	94.20	<1.00e-02	6.75e+03	6.80e+03	6.80e+03	3.66	n/a	49.60	n/a	
S981002713		A Chromium - ICP-Acid Digest	ug/g	91.20	<5.00e-02	6.75e+03	6.80e+03	6.80e+03	1.91	n/a	49.60	n/a	
S981002713		A Iron - ICP-Acid Digest	ug/g										

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002713	A	Potassium -ICP-Acid Digest	ug/g	96.20	<5.00e-01	1.28e+03	1.46e+03	1.37e+03	13.1	n/a	496.0	n/a
S98T002713	A	Lithium -ICP-Acid Digest	ug/g	97.60	<1.00e-02	<9.92e+02	<9.73e+02		n/a	n/a	9.900	n/a
S98T002713	A	Manganese -ICP-Acid Digest	ug/g	92.20	<1.00e-02	76.90	79.10	79.00	0.25	n/a	9.900	n/a
S98T002713	A	Sodium -ICP-Acid Digest	ug/g	129.2	1.150	2.31e+05	2.07e+05	2.19e+05	11.0	n/a	92.00	n/a
S98T002713	A	Nickel -ICP-Acid Digest	ug/g	91.80	<2.00e-02	1.66e+03	1.87e+03	1.76e+03	11.9	n/a	19.90	n/a
S98T002713	A	Silicon -ICP-Acid Digest	ug/g	85.80	2.130	1.72e+03	1.79e+03	1.75e+03	3.99	n/a	49.60	n/a
S98T002713	A	Uranium -ICP-Acid Digest	ug/g	93.40	<5.00e-01	3.66e+03	4.24e+03	3.95e+03	14.7	n/a	496.0	n/a
S98T002713	A	Zinc -ICP-Acid Digest	ug/g	84.60	<1.00e-02	29.50	32.50	31.00	9.68	n/a	9.900	n/a
S98T002713	A	Zirconium -ICP-Acid Digest	ug/g	93.00	<1.00e-02	<9.92e+02	11.00	n/a	n/a	n/a	9.900	n/a
S98T002714	V	Formate by IC-Dionex 4000/4500	ug/g	91.41	<6.20e-02	2.78e+02	253.0	265.6	9.42	n/a	255.4	n/a
S98T002714	V	Fluoride-IC-Dionex 4000/4500	ug/g	101.5	<1.20e-02	1.32e+04	7.54e+03	1.04e+04	54.6	n/a	120.0	n/a
S98T002714	V	Chloride-IC-Dionex 4000/4500	ug/g	100.3	<1.70e-02	1.48e+03	1.23e+03	1.35e+03	18.5	n/a	170.0	n/a
S98T002714	V	Nitrite-IC-Dionex 4000/4500	ug/g	85.23	1.21e-01	2.00e+04	1.84e+04	1.92e+04	8.33	n/a	868.4	n/a
S98T002714	V	Bromide by Ion Chromatograph	ug/g	92.32	<1.25e-01	<1.00e+03	<9.92e+02	n/a	n/a	n/a	1.00e+03	n/a
S98T002714	V	Nitrate by IC-Dionex 4000/4500	ug/g	105.2	<1.39e-01	6.96e+04	5.17e+04	6.07e+04	29.5	n/a	1.39e+03	n/a
S98T002714	V	Phosphate-IC-Dionex 4000/4500	ug/g	94.89	<1.20e-01	6.48e+04	1.28e+04	3.88e+04	134	n/a	1.20e+03	n/a
S98T002714	V	Sulfate by IC-Dionex 4000/4500	ug/g	93.73	<1.38e-01	2.60e+04	2.93e+04	2.77e+04	11.9	n/a	1.11e+03	n/a
S98T002714	V	Oxalate-IC-Dionex 4000/4500	ug/g	102.4	<1.05e-01	1.35e+04	1.35e+04	1.35e+04	0.00	n/a	844.2	n/a

Table 1. Data Summary Report.
BY-105 (2)

ON: L Lower Half of Segment

R#	Analyte	Unit	Standard %	Blank	Residue	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count Err%
81002587	Bulk Density of Sample	g/mL	n/a	n/a	n/a	n/a	n/a	n/a	n/a	5.00e-01	n/a
81002587	DSC Exotherm DRY, Calculated	Joules/g DRY	n/a	n/a	3.45e+02	298.7	321.8	14.3	n/a	n/a	n/a
81002593	DSC Exotherm on Perkin Elmer	Joules/g	96.14	n/a	2.47e+02	213.6	230.1	14.3	n/a	n/a	n/a
81002593	% Water by TGA on Perkin Elmer	%	93.84	n/a	3.91e+03	28.19	28.48	2.07	n/a	5.1000	n/a
81002593	TIC by APCI/Coulometry	ug/g	93.50	2.700	3.95e+03	3.95e+03	3.95e+03	1.02	108.0	5.1000	n/a
81002593	TIC by Persulfate/Coulometry	ug/g	92.43	2.700	1.63e+04	1.63e+04	1.76e+04	14.2	101.0	40.00	n/a
81002593	F Strontium-89/90 High Level	uCi/g	101.9	5.40e-02	6.28e+04	638.0	629.0	2.86	n/a	1.55e-01	8.28e-01
81002602	F Aluminum-1-CP-Fusion	ug/g	100.0	5.00e-02	2.80e+04	2.74e+04	2.76e+04	2.53	103.0	984.0	n/a
81002602	F Boron-1-CP-Fusion	ug/g	100.0	5.00e-02	9.85e+02	9.78e+02	n/a	n/a	100.0	984.0	n/a
81002602	F Barium-1-CP-Fusion	ug/g	101.2	5.00e-02	2.22e+03	2.80e+03	2.51e+03	23.1	103.0	1.97e+03	n/a
81002602	F Bismuth-1-CP-Fusion	ug/g	99.82	5.00e-02	8.65e+02	8.65e+02	8.65e+02	0.00	98.50	197.0	n/a
81002602	F Calcium-1-CP-Fusion	ug/g	99.40	5.00e-02	3.26e+04	3.26e+04	3.26e+04	2.10	98.50	984.0	n/a
81002602	F Chromium-1-CP-Fusion	ug/g	98.80	5.00e-02	3.06e+04	3.17e+04	3.12e+04	3.53	95.00	197.0	n/a
81002602	F Iron-1-CP-Fusion	ug/g	103.0	5.00e-02	1.97e+02	1.19e+02	n/a	n/a	99.50	197.0	n/a
81002602	F Lithium-1-CP-Fusion	ug/g	99.20	5.00e-02	4.34e+02	4.50	4.2	3.62	96.10	197.0	n/a
81002602	F Manganese-1-CP-Fusion	ug/g	104.2	5.00e-02	1.32e+05	1.29e+05	1.31e+05	2.30	144.0	1.97e+05	n/a
81002602	F Sodium-1-CP-Fusion	ug/g	99.66	1.21e-01	2.17e+04	2.33e+04	2.25e+04	7.11	98.40	984.0	n/a
81002602	F Nickel-1-CP-Fusion	ug/g	99.80	5.00e-02	1.04e+04	9.66e+03	1.00e+04	7.38	100.0	984.0	n/a
81002602	F Silicon-1-CP-Fusion	ug/g	99.20	5.00e-01	3.47e+04	3.65e+04	3.56e+04	5.06	99.40	9.84e+03	n/a
81002602	F Uranium-1-CP-Fusion	ug/g	99.60	5.00e-02	1.97e+02	1.19e+02	n/a	n/a	96.40	197.0	n/a
81002602	F Zinc-1-CP-Fusion	ug/g	99.40	5.00e-02	1.97e+02	1.19e+02	n/a	n/a	96.30	197.0	n/a
81002602	F Zirconium-1-CP-Fusion	ug/g	100.2	5.00e-02	2.60e+02	2.49	254.7	4.32	n/a	n/a	4.31e+02
81002602	F Cesium-137 by GEA	uCi/g	100.0	2.28e-01	44.32E-01	n/a	n/a	n/a	94.32	5.27E-01	3.40E+02
81002602	F Alpha of digested solid	ug/g	109.0	4.120e-03	44.82e-02	4.71e2	n/a	n/a	n/a	481.7	n/a
81002608	A Uranium-235 by ICP/MS Act1505159	ug/g	n/a	4.132e-04	< 48.18	< 4.71e1	n/a	n/a	n/a	481.7	n/a
81002608	A Uranium-235 by ICP/MS Act150	ug/g	110.6	4.132e-04	2.16e+02	254.0	234.8	16.2	92.12	481.7	n/a
81002608	A Uranium-235 by ICP/MS Act159	ug/g	n/a	4.132e-04	2.16e+02	254.0	n/a	n/a	n/a	64.25	n/a
81002608	A Uranium-236 by ICP/MS Act150	ug/g	91.50	4.132e-04	< 64.24	< 6.28e1	n/a	n/a	98.00	481.7	n/a
81002608	A Uranium-236 by ICP/MS Act159	ug/g	93.66	4.80e-05	3.08e+04	3.45e+04	3.26e+04	11.3	n/a	8.051	n/a
81002608	A Tc99 by ICP/MS Act1505159	ug/g	120.2	4.470e-03	< 8.030	< 7.85e02	n/a	n/a	n/a	602.3	n/a
81002608	A Sn126, ICP/MS Act150 LA-505159	ug/g	104.0	3.57e-05	< 18.75	< 1.83e1	n/a	n/a	115.9	18.75	n/a
81002608	A Neptunium-237 by ICP/MS	ug/g	104.0	3.57e-05	< 3.93	< 1.36e1	n/a	n/a	101.6	15.93	n/a
81002608	A Plutonium-239 by ICP/MS	ug/g	n/a	3.57e-05	< 14.27	< 1.40e1	n/a	n/a	n/a	14.27	n/a
81002608	A Plutonium-240 by ICP/MS	ug/g	110.4	2.78e-05	< 1.16	< 1.09e1	n/a	n/a	n/a	11.88	n/a
81002608	A Plutonium-242 by ICP/MS	ug/g	93.60	1.17e-05	< 14.88	< 1.46e1	n/a	n/a	107.2	14.88	n/a
81002608	A Pu/Am-241 by ICP/MS	ug/g	104.0	1.17e-01	2.47e+04	2.31e+04	2.39e+04	6.69	104.6	19.80	n/a
81002608	A Aluminum-1-CP-Acid Digest	ug/g	103.92	5.12e-01	59.80	36.70	48.25	47.9	85.20	19.80	n/a
81002608	A Boron-1-CP-Acid Digest	ug/g	91.00	5.00e-02	4.58e+02	398.0	428.0	14.0	78.40	19.80	n/a
81002608	A Barium-1-CP-Acid Digest	ug/g	88.20	5.00e-01	2.88e+03	2.51e+03	2.70e+03	13.7	72.00	39.60	n/a
81002608	A Bismuth-1-CP-Acid Digest	ug/g	92.60	1.16e-01	2.84e+04	2.67e+04	2.75e+04	5.82	-8.42e0	39.60	n/a
81002608	A Calcium-1-CP-Acid Digest	ug/g	89.60	5.00e-02	7.14e+02	646.0	680.0	10.0	77.40	3.960	n/a
81002608	A Chromium-1-CP-Acid Digest	ug/g	81.60	5.00e-02	2.67e+04	2.34e+04	2.50e+04	13.2	-5.30e1	19.80	n/a
81002608	A Iron-1-CP-Acid Digest	ug/g									

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002608	A	Potassium - IC-Acid Digest	ug/g	99.80	< 0.0e-01	1.22e+03	1.17e+03	1.20e+03	4.18	89.80	198.0
S98T002608	A	Lithium - IC-Acid Digest	ug/g	99.80	< 1.0e-02	< 3.96e-01	< 3.87e-01	< 3.92e-01	84.40	3.960	N/A
S98T002608	A	Permanesic - IC-Acid Digest	ug/g	87.80	< 1.0e-02	3.48e+02	3.17e+02	3.32e+02	9.32	75.00	N/A
S98T002608	A	Sodium - IC-Acid Digest	ug/g	121.6	9.39e-01	1.19e+02	1.20e+02	1.20e+02	0.84	-2.840e2	N/A
S98T002608	A	Strontium - IC-Acid Digest	ug/g	87.40	< 2.0e-02	1.16e+02	1.05e+02	1.10e+02	9.95	44.00	7.920
S98T002608	A	Barium - IC-Acid Digest	ug/g	157.4	1.100	1.29e+03	1.17e+03	1.23e+03	9.76	134.2	19.80
S98T002608	A	Silicon - IC-Acid Digest	ug/g	95.40	< 5.00e-01	3.66e+02	3.12e+02	3.39e+02	15.9	-8.230e1	198.0
S98T002608	A	Zinc - IC-Acid Digest	ug/g	78.80	< 1.00e-02	1.22e+02	1.10e+02	1.16e+02	10.3	73.20	3.960
S98T002608	A	Aluminum - IC-Acid Digest	ug/g	91.80	< 1.00e-02	9.620	8.650	9.135	10.6	23.00	3.960
S98T002614	A	Formate - IC-Dioxex 4000/4500	ug/g	101.7	< 6.20e-02	2.84e+02	330.0	307.1	15.0	111.8	72.60
S98T002614	A	Formate - IC-Dioxex 4000/4500	ug/g	108.6	< 1.20e-02	< 96.02	< 96.02	< 96.02	0.96	98.99	136.0
S98T002614	A	Chloride - IC-Dioxex 4000/4500	ug/g	104.5	< 1.70e-02	2.09e+03	2.07e+03	2.08e+03	3.67	106.8	864.3
S98T002614	A	Nitrite - IC-Dioxex 4000/4500	ug/g	101.2	< 5.04e-01	3.48e+02	3.41e+02	3.54e+02	N/A	N/A	N/A
S98T002614	A	Bromide by Ion Chromatograph	ug/g	99.83	< 1.25e-01	< 1.00e+03	< 1.03e+03	< 1.03e+03	N/A	97.95	1.00e+03
S98T002614	A	Nitrate by IC-Dioxex 4000/4500	ug/g	99.32	< 1.50e-01	7.55e+02	7.24e+02	7.40e+02	4.19	102.0	1.11e+03
S98T002614	A	Phosphate by IC-Dioxex 4000/4500	ug/g	104.0	< 1.20e-01	5.60e+03	5.05e+03	5.32e+03	22.9	103.3	960.2
S98T002614	A	Sulfate by IC-Dioxex 4000/4500	ug/g	101.6	< 1.38e-01	3.77e+03	3.48e+03	3.63e+03	8.00	99.69	1.10e+03
S98T002614	A	Oxalate - IC-Dioxex 4000/4500	ug/g	109.4	< 1.08e-01	8.71e+03	7.77e+03	8.24e+03	11.4	107.1	840.1

26-mar-1999 06:45:46
A-0002-1

CORE NUMBER: 246
SEGMENT #: 9A

Table 1. Data Summary Report.
BY-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk %	Det Limit	Count Err%
S981002588		Bulk Density of Sample	g/mL	N/A	N/A	Adhesive	33.78	295.3	28.8	N/A	5.00e-01	N/A
S981002594		BSC Exotherm DRY Calculated	Joules/g Dry	N/A	N/A	2.35e-02	227.9	198.6	28.8	N/A	N/A	N/A
S981002594		BSC Exotherm on Perkin Elmer	Joules/g	96.31	N/A	1.70e-02	55.41	32.77	3.91	N/A	N/A	N/A
S981002594		% Water by TGA on Perkin Elmer	%	98.84	N/A	52.13	5.19e+03	2.12e+03	2.93	N/A	5.000	N/A
S981002594		TIC by Acid/Coulometry	ug/g	98.50	2.700	5.04e-04	2.08e-04	2.12e+04	3.77	N/A	4.000	N/A
S981002594		TIC by Persulfate/Coulometry	ug/g	92.67	2.700	2.16e-04	810.0	789.5	5.19	N/A	1.63e-01	7.61E-01
S981002594		Fe Strontium-89/90 High Level	uCi/g	101.9	5.40e-02	7.69e-04	2.28e-04	2.26e+04	1.32	N/A	1.02e+03	N/A
S981002603		Aluminum - ICP-Fusion	ug/g	100.0	< 5.00e-02	2.23e-04	4.04e3	N/A	N/A	N/A	1.02e+03	N/A
S981002603		Boron - ICP-Fusion	ug/g	101.2	< 5.00e-02	1.03e-03	4.10e3	N/A	N/A	N/A	1.02e+03	N/A
S981002603		Bismuth - ICP-Fusion	ug/g	101.2	< 5.00e-02	3.51e-03	3.25e+03	3.28e+03	1.83	N/A	2.05e+03	N/A
S981002603		Barium - ICP-Fusion	ug/g	99.80	< 1.00e-01	2.23e-04	2.31e+04	2.28e+04	2.63	N/A	2.05e+03	N/A
S981002603		Calcium - ICP-Fusion	ug/g	99.40	< 1.00e-01	4.65e-02	1.32e0	443.5	9.70	N/A	205.0	N/A
S981002603		Chromium - ICP-Fusion	ug/g	98.80	< 5.00e-02	3.48e-04	3.53e+04	3.50e+04	1.43	N/A	1.02e+03	N/A
S981002603		Lithium - ICP-Fusion	ug/g	103.0	< 1.00e-02	4.67e-02	4.55e0	N/A	N/A	N/A	205.0	N/A
S981002603		Manganese - ICP-Fusion	ug/g	99.20	< 1.00e-02	4.67e-02	1.45e+03	1.42e+05	4.23	N/A	2.05e+03	N/A
S981002603		Sodium - ICP-Fusion	ug/g	106.2	1.21e-01	2.37e-04	2.28e+04	2.42e+04	12.0	N/A	410.0	N/A
S981002603		Nickel - ICP-Fusion	ug/g	99.60	< 5.00e-02	6.19e-03	5.97e+03	6.08e+03	3.62	N/A	1.02e+03	N/A
S981002603		Silicon - ICP-Fusion	ug/g	99.20	< 5.00e-01	4.01e-04	4.07e+04	4.04e+04	1.49	N/A	205.0	N/A
S981002603		Uranium - ICP-Fusion	ug/g	95.40	< 1.00e-02	4.03e-02	2.09e2	N/A	N/A	N/A	205.0	N/A
S981002603		Zirconium - ICP-Fusion	ug/g	98.40	< 1.00e-02	4.03e-02	2.09e2	N/A	N/A	N/A	205.0	N/A
S981002603		Zirconium-137 by GEA	uCi/g	100.2	N/A	2.17e-02	20.0	218.6	1.37	N/A	5.49e-01	1.41E+02
S981002603		Alpha of Digested Solid	uCi/g	100.0	< 2.28e-01	< 3.94e-01	4.90e+01	N/A	N/A	N/A	495.5	N/A
S981002603		A 1120 by ICP/MS Acid/d1A505159 ug/g	ug/g	109.0	< 1.20e-03	4.56e-02	44.65e2	N/A	N/A	N/A	49.55	N/A
S981002609		Uranium-233 by ICP/MS Acid/d159 ug/g	ug/g	N/A	< 1.32e-04	< 49.33	44.65e3	N/A	N/A	N/A	49.55	N/A
S981002609		Uranium-234 by ICP/MS Acid/d159 ug/g	ug/g	110.6	< 1.32e-04	3.19e-02	287.0	302.0	10.6	N/A	49.55	N/A
S981002609		Uranium-235 by ICP/MS Acid/d159 ug/g	ug/g	N/A	< 1.76e-04	< 66.30	3.89e+04	4.15e+04	12.8	N/A	66.08	N/A
S981002609		Uranium-236 by ICP/MS Acid/d159 ug/g	ug/g	91.50	< 1.32e-04	4.42e-04	27.76e0	N/A	N/A	N/A	49.55	N/A
S981002609		Uranium-238 by ICP/MS Acid/d159 ug/g	ug/g	93.66	< 1.32e-04	4.42e-04	27.76e0	N/A	N/A	N/A	49.55	N/A
S981002609		Uranium-238 by ICP/MS Acid/d1A505159 ug/g	ug/g	120.2	< 4.67e-05	< 19.62	< 8.25e2	N/A	N/A	N/A	619.4	N/A
S981002609		Strontium-87 by ICP/MS	ug/g	N/A	< 1.50e-03	< 1.92e-02	< 8.25e2	N/A	N/A	N/A	19.28	N/A
S981002609		Plutonium-239 by ICP/MS	ug/g	104.0	< 3.47e-05	< 14.33	< 1.35e1	N/A	N/A	N/A	14.33	N/A
S981002609		Plutonium-240 by ICP/MS	ug/g	N/A	< 3.55e-05	< 14.60	< 1.38e1	N/A	N/A	N/A	14.67	N/A
S981002609		Plutonium-242 by ICP/MS	ug/g	N/A	< 2.78e-05	< 15.30	< 1.44e1	N/A	N/A	N/A	15.30	N/A
S981002609		Plutonium-241 by ICP/MS	ug/g	110.1	< 3.71e-05	< 15.30	< 1.44e1	N/A	N/A	N/A	15.30	N/A
S981002609		Aluminum - ICP-Acid Digest	ug/g	93.60	1.17e-01	2.07e-04	1.89e+04	1.98e+04	9.09	N/A	20.40	N/A
S981002609		Boron - ICP-Acid Digest	ug/g	104.2	5.42e-01	80.00	82.00	472.0	12.3	N/A	20.40	N/A
S981002609		Barium - ICP-Acid Digest	ug/g	94.00	< 5.00e-02	5.01e-02	443.0	N/A	N/A	N/A	40.60	N/A
S981002609		Bismuth - ICP-Acid Digest	ug/g	88.20	< 1.00e-01	3.68e-03	3.31e+03	3.47e+03	9.22	N/A	40.60	N/A
S981002609		Barium - ICP-Acid Digest	ug/g	90.40	1.46e-01	2.19e-04	2.01e+04	2.10e+04	8.57	N/A	4.060	N/A
S981002609		Calcium - ICP-Acid Digest	ug/g	89.60	< 1.00e-02	3.87e-02	3.65e0	3.66e5	11.2	N/A	20.40	N/A
S981002609		Chromium - ICP-Acid Digest	ug/g	87.60	< 5.00e-02	3.13e-04	2.85e+04	2.99e+04	9.36	N/A	20.40	N/A
S981002609		Iron - ICP-Acid Digest	ug/g	87.60	< 5.00e-02	3.13e-04	2.85e+04	2.99e+04	9.36	N/A	20.40	N/A

Sample#	P#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002609	A	Potassium -ICP-Acid Digest	ug/g	91.80	< 1.00e-01	9.00e+03	< 855.0	877.5	5.13	n/a	204.0	n/a
S98T002609	A	Lithium -ICP-Acid Digest	ug/g	99.80	< 1.00e-02	3.5e+02	< 352.0	n/a	n/a	n/a	4.060	n/a
S98T002609	A	Manganese -ICP-Acid Digest	ug/g	87.80	< 1.00e-02	3.80e+02	352.0	349.0	9.28	n/a	4.060	n/a
S98T002609	A	Sodium -ICP-Acid Digest	ug/g	121.0	9.39e-01	3.90e+02	1.20e+03	1.30e+03	14.7	n/a	40.40	n/a
S98T002609	A	Strontium -ICP-Acid Digest	ug/g	87.40	< 2.00e-02	1.27e+02	1.17e+03	1.18e+03	10.2	n/a	8.140	n/a
S98T002609	A	Barium -ICP-Acid Digest	ug/g	157.4	< 1.00e-01	2.45e+02	1.72e+03	2.08e+03	35.0	n/a	20.40	n/a
S98T002609	A	Silicon -ICP-Acid Digest	ug/g	95.40	< 5.00e-01	6.35e+02	3.62e+04	4.08e+04	13.0	n/a	204.0	n/a
S98T002609	A	Zinc -ICP-Acid Digest	ug/g	78.80	< 1.00e-02	1.43e+02	< 352.0	135.0	14.8	n/a	4.060	n/a
S98T002609	A	Titanium -ICP-Acid Digest	ug/g	91.80	< 1.00e-02	3.13e+02	< 352.0	n/a	n/a	n/a	4.060	n/a
S98T002609	A	Fluoride by IC-Dionex 4000/4500	ug/g	101.7	< 6.20e-02	3.60e+02	353.0	356.0	1.93	n/a	74.46	n/a
S98T002615	W	Fluoride -IC-Dionex 4000/4500	ug/g	105.4	< 1.20e-02	3.18e+02	< 9.563	n/a	n/a	n/a	38.48	n/a
S98T002615	W	Chloride -IC-Dionex 4000/4500	ug/g	99.87	< 1.70e-02	1.60e+02	1.71e+03	1.65e+03	6.76	n/a	139.5	n/a
S98T002615	W	Nitrate by IC-Dionex 4000/4500	ug/g	100.0	2.43e-01	3.07e+02	3.79e+04	3.18e+04	6.92	n/a	886.4	n/a
S98T002615	W	Nitrate by IC-Dionex 4000/4500	ug/g	97.44	< 1.50e-01	< 1.03e+02	< 95e+2	n/a	n/a	n/a	1.03e+03	n/a
S98T002615	W	Nitrate by IC-Dionex 4000/4500	ug/g	94.76	< 1.50e-01	5.88e+02	5.33e+04	5.60e+04	9.81	n/a	1.14e+03	n/a
S98T002615	W	Phosphate by IC-Dionex 4000/4500	ug/g	100.2	< 1.20e-01	2.48e+02	1.68e+04	2.08e+04	38.5	n/a	984.8	n/a
S98T002615	W	Sulfate by IC-Dionex 4000/4500	ug/g	100.8	< 1.30e-01	2.22e+03	3.21e+03	2.72e+03	36.5	n/a	1.13e+03	n/a
S98T002615	W	Oxalate by IC-Dionex 4000/4500	ug/g	109.4	< 1.00e-01	1.47e+04	1.50e+04	1.43e+04	6.32	n/a	861.8	n/a

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 246
SEGMENT #: FIELD BLANK

SEGMENT PORTION: Field Blank

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002391		Chromium (VI) by Spec.	ug/mL	100.8	<2.58e-01	<0.141	<0.141	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		DSC Exotherm DRY, Calculated	Joules/g Dry	N/A	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	1.00e-02	N/A
S98T002391		DSC Exotherm on Perkin Elmer	Joules/g	100.4	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	1.00e-02	N/A
S98T002391		D I 129 by ICP/MS Acid Addition	ug/mL	104.7	<1.20e-04	<1.21e-02	<1.21e-02	35.10	24.8	N/A	1.20e-02	N/A
S98T002391		Ammonia by ISE-std Additions	ug/mL	90.76	<10.00	29.70	36.10	33.50	24.8	N/A	63.00	N/A
S98T002391		Oil by Pot. Titration	ug/mL	100.0	N/A	< 63.00	<63	99.31	0.12	N/A	63.00	N/A
S98T002391		% Water by TGA on Perkin Elmer	%	98.69	N/A	< 40.00	<40.00	99.31	0.12	N/A	40.00	N/A
S98T002391		TOC by Persulfate/Oxymetry	ug/mL	99.33	<4.000	< 40.00	<4.00e+1	99.31	0.12	N/A	40.00	N/A
S98T002391		Strontium-89/90 High Level	ug/mL	99.04	<2.08e-07	<3.23e-07	<4.13E-7	N/A	N/A	N/A	4.35e-07	3.11E+02
S98T002391		Pu-239/240 by TRU-SPEC Resin	ug/mL	113.4	<6.56e-07	<6.71e-07	<6.71E-7	N/A	N/A	N/A	6.71e-07	1.00E+02
S98T002391		D Uranium-235 by ICP/MS Acid Add	ug/mL	N/A	<1.20e-05	<1.21e-03	<1.21e-3	N/A	N/A	N/A	1.00e-03	N/A
S98T002391		D Uranium-234 by ICP/MS Acid Add	ug/mL	N/A	<1.20e-05	<1.21e-03	<1.21e-3	N/A	N/A	N/A	1.00e-03	N/A
S98T002391		D Uranium-235 by ICP/MS Acid Add	ug/mL	102.8	<1.20e-05	<1.21e-03	<1.21e-3	N/A	N/A	N/A	1.00e-03	N/A
S98T002391		D Uranium-236 by ICP/MS Acid Add	ug/mL	102.5	<1.20e-05	<1.21e-03	<1.21e-3	N/A	N/A	N/A	1.00e-03	N/A
S98T002391		D Uranium-238 by ICP/MS Acid Add	ug/mL	102.5	<1.20e-05	<1.21e-03	<1.21e-3	N/A	N/A	N/A	1.00e-03	N/A
S98T002391		D Tc 99 by ICP/MS Acid Addition	ug/mL	89.37	<2.00e-06	<2.20e-03	<2.20E-3	N/A	N/A	N/A	2.20e-03	N/A
S98T002391		D Sn 126 by ICP/MS Acid Addition	ug/mL	N/A	<1.35e-04	<1.60e-03	<1.60E-3	N/A	N/A	N/A	2.00e-03	N/A
S98T002391		D Thorium-229 by ICP/MS	ug/mL	N/A	<3.68e-06	<4.15e-04	<4.15E-4	N/A	N/A	N/A	3.70e-04	N/A
S98T002391		D Neptunium-237 by ICP/MS	ug/mL	124.3	<6.30e-06	<6.15e-04	<6.15E-4	N/A	N/A	N/A	6.15e-04	N/A
S98T002391		D Plutonium-239 by ICP/MS	ug/mL	109.0	<6.07e-06	<5.95e-04	<5.95E-4	N/A	N/A	N/A	6.13e-04	N/A
S98T002391		D Plutonium-240 by ICP/MS	ug/mL	N/A	<2.24e-03	<2.7E-04	<2.7E-4	N/A	N/A	N/A	2.72e-04	N/A
S98T002391		D Plutonium-242 by ICP/MS	ug/mL	N/A	<6.68e-06	<6.42e-04	<6.42E-4	N/A	N/A	N/A	6.42e-04	N/A
S98T002391		D Pu/Am-241 by ICP/MS	ug/mL	112.6	<6.30e-06	<6.15e-04	<6.15E-4	N/A	N/A	N/A	5.47e-04	N/A
S98T002391		D Aluminum-ICP-Acid DIL.	ug/mL	98.20	<5.00e-02	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002391		D Beryllium-ICP-Acid DIL.	ug/mL	99.00	<5.00e-02	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002391		D Boron-ICP-Acid DIL.	ug/mL	98.60	<5.00e-02	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002391		D Barium-ICP-Acid DIL.	ug/mL	100.6	<5.00e-02	<5.00e-02	<5.00e-2	N/A	N/A	N/A	1.00e-01	N/A
S98T002391		D Bismuth-ICP-Acid DIL.	ug/mL	97.20	<1.00e-01	<2.33e-01	<2.19e-01	2.26e-01	6.19	N/A	1.00e-01	N/A
S98T002391		D Calcium-ICP-Acid DIL.	ug/mL	100.2	<1.00e-01	<4.00e-02	<4.00e-2	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		D Chromium-ICP-Acid DIL.	ug/mL	98.60	<5.00e-02	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002391		D Iron-ICP-Acid DIL.	ug/mL	98.60	<5.00e-01	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S98T002391		D Potassium-ICP-Acid DIL.	ug/mL	98.60	<1.00e-02	<1.00e-02	<1.00e-2	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		D Lithium-ICP-Acid DIL.	ug/mL	98.60	<1.00e-02	<1.00e-02	<1.00e-2	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		D Manganese-ICP-Acid DIL.	ug/mL	103.8	<1.00e-01	<1.23e-01	<1.25e-01	1.240	1.61	N/A	1.00e-01	N/A
S98T002391		D Sodium-ICP-Acid DIL.	ug/mL	107.2	<2.00e-02	<2.00e-02	<2.00e-2	N/A	N/A	N/A	2.00e-02	N/A
S98T002391		D Nickel-ICP-Acid DIL.	ug/mL	98.20	<5.00e-02	<9.93e-01	<9.84e-01	9.88e-01	0.91	N/A	5.00e-02	N/A
S98T002391		D Silicon-ICP-Acid DIL.	ug/mL	98.20	<1.00e-01	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S98T002391		D Uranium-ICP-Acid DIL.	ug/mL	98.20	<1.00e-02	<1.00e-02	<1.00e-2	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		D Zinc-ICP-Acid DIL.	ug/mL	98.00	<1.00e-02	<1.00e-02	<1.00e-2	N/A	N/A	N/A	1.00e-02	N/A
S98T002391		D Zirconium-ICP-Acid DIL.	ug/mL	97.08	<2.00e-02	<4.20e-02	<4.20e-2	N/A	N/A	N/A	1.20e-02	N/A
S98T002391		Formate by Ion Chromatograph	ug/mL	97.84	<2.00e-02	<4.20e-02	<4.20e-2	N/A	N/A	N/A	1.20e-02	N/A
S98T002391		Fluoride-IC-Dionex 4000/4500	ug/mL	97.78	<2.00e-02	<4.20e-02	<4.20e-2	3.30e-02	0.00	N/A	1.70e-02	N/A
S98T002391		Chloride-IC-Dionex 4000/4500	ug/mL	97.78	<2.00e-02	<4.20e-02	<4.20e-2	3.30e-02	0.00	N/A	1.70e-02	N/A
S98T002391		Nitrite-IC - Dionex 4000/4500	ug/mL	97.35	<1.00e-01	<1.08e-01	<1.08e-1	N/A	N/A	N/A	1.08e-01	N/A
S98T002391		Bromide by Ion Chromatograph	ug/mL	95.71	<1.25e-01	<1.25e-01	<1.25e-1	N/A	N/A	N/A	1.25e-01	N/A

26-mar-1999 06:49:39
K-0002-1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002391		Nitrate by IC-Dionex 4000/4500	ug/mL	97.71	<1.39e-01	<1.39e-01	<1.39e-01	N/A	N/A	N/A	1.39e-01	N/A
S98T002391		Phosphate-IC-Dionex 4000/4500	ug/mL	91.61	<1.20e-01	<1.20e-01	<1.20e-01	N/A	N/A	N/A	1.20e-01	N/A
S98T002391		Sulfate by IC-Dionex 4000/4500	ug/mL	95.56	<1.38e-01	<1.38e-01	<1.38e-01	N/A	N/A	N/A	1.38e-01	N/A
S98T002391		Oxalate-IC-Dionex 4000/4500	ug/mL	96.11	<1.05e-01	<1.05e-01	<1.05e-01	N/A	N/A	N/A	1.05e-01	N/A
S98T002391		Cesium-137 by GEA	uCi/mL	103.6	<3.30e-03	<3.30e-03	<3.30e-03	N/A	N/A	N/A	3.34e-05	N/A
S98T002391		Alpha in Liquid Samples	uCi/mL	89.27	<3.78e-08	<7.02e-08	<6.26e-8	N/A	N/A	N/A	1.43e-07	5.00E+02

Table 1. Data Summary Report.
BY-105 (2)

CORE NUMBER: 246
SEGMENT #: n/a

SEGMENT PORTION: Hydrostatic Head Fluid

Sample#	R #	Analyte	Unit	Standard X	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002392	D	Aluminum-ICP-Acid Dil.	ug/mL	98.20	< 5.00e-02	< 5.00e-01	< 5.00e-1	n/a	n/a	n/a	2.00e-01	n/a
S98T002392	D	Boron-ICP-Acid Dil.	ug/mL	99.00	< 5.00e-02	< 5.00e-01	< 5.00e-1	n/a	n/a	n/a	2.00e-01	n/a
S98T002392	D	Barium-ICP-Acid Dil.	ug/mL	98.60	< 5.00e-02	< 5.00e-01	< 5.00e-1	n/a	n/a	n/a	2.00e-01	n/a
S98T002392	D	Bismuth-ICP-Acid Dil.	ug/mL	100.60	< 1.00e-01	< 1.000	< 1.00e0	n/a	n/a	n/a	1.000	n/a
S98T002392	D	Calcium-ICP-Acid Dil.	ug/mL	97.60	< 1.00e-01	< 1.000	< 1.00e0	n/a	n/a	n/a	1.000	n/a
S98T002392	D	Chromium-ICP-Acid Dil.	ug/mL	100.2	< 1.00e-02	< 1.00e-01	< 1.00e-1	n/a	n/a	n/a	1.00e-01	n/a
S98T002392	D	Iron-ICP-Acid Dil.	ug/mL	96.60	< 5.00e-02	< 5.00e-01	< 5.00e-1	n/a	n/a	n/a	5.00e-01	n/a
S98T002392	D	Potassium-ICP-Acid Dil.	ug/mL	96.80	< 5.00e-01	< 5.000	< 5.00e0	n/a	n/a	n/a	5.000	n/a
S98T002392	D	Lithium-ICP-Acid Dil.	ug/mL	98.60	< 1.00e-02	1.88e-03	1.89e-03	1.88e-03	0.53	n/a	1.00e-01	n/a
S98T002392	D	Manganese-ICP-Acid Dil.	ug/mL	98.80	< 1.00e-02	6.110	5.960	6.045	2.15	n/a	1.000	n/a
S98T002392	D	Sodium-ICP-Acid Dil.	ug/mL	102.8	< 1.00e-01	< 2.00e-01	< 2.00e-1	n/a	n/a	n/a	2.00e-01	n/a
S98T002392	D	Nickel-ICP-Acid Dil.	ug/mL	101.2	< 2.00e-02	< 3.180	3.140	3.160	1.21	n/a	5.00e-01	n/a
S98T002392	D	Silicon-ICP-Acid Dil.	ug/mL	96.20	< 5.00e-02	< 5.000	< 5.00e0	n/a	n/a	n/a	5.000	n/a
S98T002392	D	Uranium-ICP-Acid Dil.	ug/mL	94.40	< 1.00e-02	< 1.00e-01	< 1.00e-1	n/a	n/a	n/a	1.00e-01	n/a
S98T002392	D	Zinc-ICP-Acid Dil.	ug/mL	99.00	< 1.00e-02	< 1.00e-01	< 1.00e-1	n/a	n/a	n/a	1.00e-01	n/a
S98T002392	D	Fluoride-IC-Dionex 4000/4500	ug/mL	94.84	< 1.20e-02	< 25.43	< 25.61	n/a	n/a	n/a	25.45	n/a
S98T002392	D	Chloride-IC-Dionex 4000/4500	ug/mL	97.78	< 1.70e-02	< 36.06	< 36.43	n/a	n/a	n/a	36.06	n/a
S98T002392	D	Nitrite-IC - Dionex 4000/4500	ug/mL	94.35	< 1.00e-01	< 2.29e-02	< 2.30e-2	2.41e-04	1.20	n/a	229.1	n/a
S98T002392	D	Bromide by Ion Chromatograph	ug/mL	95.71	< 1.20e-01	< 2.68e-02	< 2.69e-2	n/a	n/a	n/a	265.1	n/a
S98T002392	D	Nitrate by IC-Dionex 4000/4500	ug/mL	97.61	< 1.20e-01	< 2.68e-02	< 2.69e-2	n/a	n/a	n/a	254.8	n/a
S98T002392	D	Phosphate-IC-Dionex 4000/4500	ug/mL	97.61	< 1.20e-01	< 2.68e-02	< 2.69e-2	n/a	n/a	n/a	254.5	n/a
S98T002392	D	Sulfate by IC-Dionex 4000/4500	ug/mL	95.56	< 1.20e-01	< 2.68e-02	< 2.69e-2	n/a	n/a	n/a	292.7	n/a
S98T002392	D	Oxalate-IC-Dionex 4000/4500	ug/mL	96.11	< 1.00e-01	< 2.26e-02	< 2.35e-2	n/a	n/a	n/a	222.7	n/a

Table 1. Data Summary Report.
BY-105 (2)

CORE NUMBER: 246

SEGMENT #: Core Composite

SEGMENT PORTION: Core Composite

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003375		Bulk Density of Sample	g/mL	N/A	N/A	DRY	N/A	N/A	N/A	N/A	5.00E-01	N/A
S98T003375		Bulk Density of Sample	g/mL	N/A	N/A	DRY	N/A	N/A	N/A	N/A	5.00E-01	N/A
S98T003391		DSC Exotherm DRY Calculated	Joules/g DRY	N/A	N/A	11.82	21.76	16.79	59.2	N/A	N/A	N/A
S98T003391		DSC Exotherm on Perkin Elmer	Joules/g	96.66	N/A	11.82	21.24	16.39	59.2	N/A	N/A	N/A
S98T003391		% Water by TGA on Perkin Elmer	%	102.1	N/A	2.530	2.210	2.370	13.5	88.80	5.000	N/A
S98T003391		TIC by Acid/Coulometry	ug/g	90.36	4.500	6.17E+03	5.20E+03	5.68E+03	17.1	94.30	40.00	N/A
S98T003391		TOC by Persulfate/Coulometry	ug/g	93.33	15.10	6.34E+02	656.0	645.0	3.41	94.30	40.00	N/A
S98T003393		F Strontium-89/90 High Level	uCi/g	99.34	3.80E-02	1.010	8.55E-01	9.32E-01	16.6	N/A	2.00E-03	1.46E+00
S98T003393		F Aluminum - ICP-Fusion	ug/g	97.40	<5.00E-02	1.02E+03	8.76E+03	8.56E+03	4.67	97.00	1.02E+03	N/A
S98T003393		F Boron - ICP-Fusion	ug/g	107.8	<5.00E-02	<1.0E+03	<1.0E+03	N/A	N/A	105.0	1.02E+03	N/A
S98T003393		F Barium - ICP-Fusion	ug/g	96.40	<5.00E-02	<1.0E+03	<1.0E+03	N/A	N/A	92.30	1.02E+03	N/A
S98T003393		F Bismuth - ICP-Fusion	ug/g	100.4	<1.00E-01	<2.0E+03	<2.0E+03	N/A	N/A	95.10	2.04E+03	N/A
S98T003393		F Calcium - ICP-Fusion	ug/g	101.8	<1.00E-01	<2.0E+03	<2.0E+03	N/A	N/A	99.20	2.04E+03	N/A
S98T003393		F Chromium - ICP-Fusion	ug/g	99.60	<1.00E-02	<2.0E+03	<2.0E+03	N/A	N/A	99.00	204.0	N/A
S98T003393		F Iron - ICP-Fusion	ug/g	95.80	<5.00E-02	<1.0E+03	<1.0E+03	N/A	N/A	95.00	1.02E+03	N/A
S98T003393		F Lithium - ICP-Fusion	ug/g	99.80	<1.00E-02	<2.0E+03	<2.0E+03	N/A	N/A	95.00	204.0	N/A
S98T003393		F Manganese - ICP-Fusion	ug/g	99.20	<1.00E-02	2.45E+02	2.50E+02	2.48E+02	2.02	95.40	204.0	N/A
S98T003393		F Nickel - ICP-Fusion	ug/g	98.80	7.40E-02	9.16E+02	596.0	756.0	42.3	97.40	407.0	N/A
S98T003393		F Silicon - ICP-Fusion	ug/g	107.4	<5.00E-02	<1.0E+03	<1.0E+03	N/A	N/A	112.0	1.02E+03	N/A
S98T003393		F Uranium - ICP-Fusion	ug/g	98.20	<5.00E-01	<1.0E+04	<1.0E+04	N/A	N/A	93.00	1.02E+04	N/A
S98T003393		F Zinc - ICP-Fusion	ug/g	96.20	<1.00E-02	<2.0E+02	<2.0E+02	N/A	N/A	96.60	204.0	N/A
S98T003393		F Zirconium - ICP-Fusion	ug/g	101.4	<1.00E-02	<2.0E+02	<2.0E+02	N/A	N/A	99.10	204.0	N/A
S98T003393		F Cesium-137 by BEA	uCi/g	97.41	<1.00E-02	8.342	8.860	8.601	6.05	N/A	0.970	N/A
S98T003393		F Alpha of Digested Solid	uCi/g	93.87	<1.50E-01	<1.0E-01	<1.09E-1	N/A	N/A	89.88	2.01E-01	3.03E+02
S98T003394		A 1129 by ICP/MS Acid/LA505159	ug/g	109.0	<1.20E-03	<4.74E-01	<4.96E-1	N/A	N/A	N/A	4.740	N/A
S98T003394		A Uranium-233 by ICP/MS Acid/LA505159	ug/g	N/A	<1.20E-04	44.74E-01	44.96E-1	N/A	N/A	N/A	4.74E-01	N/A
S98T003394		A Uranium-234 by ICP/MS Acid/LA505159	ug/g	N/A	<1.20E-04	44.74E-01	44.96E-1	N/A	N/A	N/A	4.74E-01	N/A
S98T003394		A Uranium-235 by ICP/MS Acid/LA505159	ug/g	102.8	<1.20E-04	44.74E-01	44.96E-1	N/A	N/A	105.3	4.74E-01	N/A
S98T003394		A Uranium-236 by ICP/MS Acid/LA505159	ug/g	N/A	<1.20E-04	44.74E-01	44.96E-1	N/A	N/A	N/A	6.32E-01	N/A
S98T003394		A Uranium-238 by ICP/MS Acid/LA505159	ug/g	93.66	<1.20E-04	20.85	25.70	23.27	20.6	104.5	4.74E-01	N/A
S98T003394		A Tc99 by ICP/MS Acid/LA505159	ug/g	103.65	1.47E-04	2.69E-01	3.10E-01	2.89E-01	14.2	N/A	7.90E-02	N/A
S98T003394		A Sr126 ICP/MS Acid/LA505159	ug/g	N/A	<1.50E-03	<5.92E-01	<6.20E-1	N/A	N/A	N/A	5.92E-01	N/A
S98T003394		A Neptunium-237 by ICP/MS	ug/g	119.4	44.67E-03	1.84E-01	1.93E-1	N/A	N/A	117.4	1.84E-01	N/A
S98T003394		A Plutonium-239 by ICP/MS	ug/g	102.4	3.47E-05	1.37E-01	1.43E-1	N/A	N/A	102.1	1.37E-01	N/A
S98T003394		A Plutonium-234 by ICP/MS	ug/g	N/A	<3.55E-05	1.40E-01	1.47E-1	N/A	N/A	N/A	1.40E-01	N/A
S98T003394		A Plutonium-242 by ICP/MS	ug/g	N/A	<2.78E-05	1.15E-01	1.15E-1	N/A	N/A	N/A	1.15E-01	N/A
S98T003394		A Pu/Am-241 by ICP/MS	ug/g	107.8	<3.77E-05	1.46E-01	1.53E-1	N/A	N/A	106.2	1.46E-01	N/A
S98T003394		A Aluminum - ICP-Acid Digest	ug/g	94.60	1.32E-01	7.89E+03	9.51E+03	8.70E+03	18.6	132.8	49.10	N/A
S98T003394		A Boron - ICP-Acid Digest	ug/g	106.0	7.61E-01	1.43E+02	157.0	150.0	9.33	101.0	49.10	N/A
S98T003394		A Barium - ICP-Acid Digest	ug/g	93.60	<5.00E-02	<49.10	<5.14E1	N/A	N/A	95.40	49.10	N/A
S98T003394		A Bismuth - ICP-Acid Digest	ug/g	89.80	<1.00E-01	<98.30	<1.03E2	N/A	N/A	99.20	98.30	N/A
S98T003394		A Calcium - ICP-Acid Digest	ug/g	97.60	1.65E-01	<98.30	<1.03E2	N/A	N/A	114.8	98.30	N/A
S98T003394		A Chromium - ICP-Acid Digest	ug/g	94.00	<1.00E-02	1.24E+02	145.0	134.5	15.6	101.6	9.850	N/A

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981003394	A	Iron - ICP-Acid Digest	ug/g	91.40	<5.00e-02	1.98e+03	112.0	1.05e+03	n/a	9.62e01	49.10	n/a
S981003394	A	Potassium - ICP-Acid Digest	ug/g	90.60	<5.00e-01	<4.91e+02	<5.14e2	n/a	n/a	110.2	492.0	n/a
S981003394	A	Lithium - ICP-Acid Digest	ug/g	95.60	<1.00e-02	<9.83e-01	<1.03e1	n/a	n/a	97.20	9.850	n/a
S981003394	A	Manganese - ICP-Acid Digest	ug/g	92.00	<1.00e-02	15.00	<1.03e1	n/a	n/a	96.40	9.850	n/a
S981003394	A	Sodium - ICP-Acid Digest	ug/g	113.8	1.180	2.39e+05	2.49e+05	2.44e+05	4.10	360.0	98.50	n/a
S981003394	A	Nickel - ICP-Acid Digest	ug/g	91.60	<2.00e-02	<19.70	<2.06e1	n/a	n/a	101.0	19.70	n/a
S981003394	A	Silicon - ICP-Acid Digest	ug/g	163.6	2.230	3.01e+02	3.75e1	337.0	21.4	154.2	49.10	n/a
S981003394	A	Uranium - ICP-Acid Digest	ug/g	92.90	<1.00e-01	<4.91e+02	<5.14e2	n/a	n/a	95.40	492.0	n/a
S981003394	A	Zinc - ICP-Acid Digest	ug/g	83.60	<1.00e-02	<9.83e-01	<1.03e1	n/a	n/a	97.60	9.850	n/a
S981003394	A	Zirconium - ICP-Acid Digest	ug/g	92.40	<1.00e-02	<9.83e-01	<1.03e1	n/a	n/a	97.60	9.850	n/a
S981003394	A	Chromium (VI) by Spec.	ug/g	100.8	<27.9e	1.03e+02	101.0	102.0	1.96	100.5	1.00e-02	n/a
S981003395	W	OH- by Pot. Titration	ug/g	96.51	<24.82e-01	<2.33e+04	<24234	n/a	n/a	n/a	2.53e+04	n/a
S981003395	W	Formate by IC-Dioxex 4000/4500	ug/g	95.80	<6.20e-02	<1.38e+02	<1.32e2	n/a	n/a	94.96	138.1	n/a
S981003395	W	Fluoride-IC-Dioxex 4000/4500	ug/g	100.5	<1.20e-02	<4.10e+02	<3.72e1	391.2	9.72	96.62	245.2	n/a
S981003395	W	Chloride-IC-Dioxex 4000/4500	ug/g	97.86	<1.70e-02	<4.02e+02	<3.35e2	n/a	n/a	93.47	347.4	n/a
S981003395	W	Nitrite-IC - Dioxex 4000/4500	ug/g	94.37	1.24e-01	3.17e+03	3.13e+03	3.15e+03	1.27	89.71	2.21e+03	n/a
S981003395	W	Bromide by Ion Chromatograph	ug/g	95.22	<1.25e-01	<2.55e+03	<2.45e3	n/a	n/a	89.25	2.55e+03	n/a
S981003395	W	Nitrate by IC-Dioxex 4000/4500	ug/g	95.95	<1.39e-01	<6.55e+05	<6.72e+05	6.64e+05	2.56	70.61	2.84e+03	n/a
S981003395	W	Phosphate-IC-Dioxex 4000/4500	ug/g	98.91	<1.20e-01	<2.45e+03	<2.35e3	n/a	n/a	88.69	2.45e+03	n/a
S981003395	W	Sulfate by IC-Dioxex 4000/4500	ug/g	96.39	<1.38e-01	<2.82e+03	<2.70e3	n/a	n/a	94.04	2.82e+03	n/a
S981003395	W	Oxalate-IC-Dioxex 4000/4500	ug/g	104.4	<1.05e-02	<2.16e+03	<2.06e3	n/a	n/a	97.13	2.15e+03	n/a
S981003396		DSC Exotherm Dry Calculated	Joules/g	n/a	n/a	1.81e+02	187.8	184.4	3.68	n/a	n/a	n/a
S981003396		% Water by TGA on Perkin Elmer	%	96.66	n/a	1.26e+02	130.3	128.0	3.67	n/a	n/a	n/a
S981003396		TIC by Acid/Coulometry	ug/g	102.1	n/a	31.28	29.96	30.62	4.31	n/a	n/a	n/a
S981003396		TOC by Persulfate/Coulometry	ug/g	90.36	4.500	4.30e+03	5.10e+03	4.70e+03	17.0	103.0	5.000	n/a
S981003398	F	Strontium-87/90 High Level	uCi/g	93.33	15.10	2.19e+04	2.09e+04	2.14e+04	4.67	74.60	40.00	n/a
S981003398	F	Aluminum - ICP-Fusion	ug/g	99.04	<3.80e-02	6.84e+04	689.0	686.5	0.73	n/a	2.05e-01	5.64e-01
S981003398	F	Boron - ICP-Fusion	ug/g	97.40	<5.00e-02	2.51e+04	2.47e+04	2.49e+04	1.61	93.60	1.03e+03	n/a
S981003398	F	Barium - ICP-Fusion	ug/g	107.8	<5.00e-02	<1.03e+03	<1.03e3	n/a	n/a	103.0	1.03e+03	n/a
S981003398	F	Bismuth - ICP-Fusion	ug/g	96.40	<5.00e-01	<2.93e+03	<3.07e+03	3.00e+03	4.67	100.0	2.07e+03	n/a
S981003398	F	Calcium - ICP-Fusion	ug/g	101.8	<1.00e-01	2.75e+04	2.83e+04	2.78e+04	3.60	104.0	2.07e+03	n/a
S981003398	F	Chromium - ICP-Fusion	ug/g	99.60	<1.00e-02	2.06e+02	652.0	679.0	7.95	98.80	207.0	n/a
S981003398	F	Iron - ICP-Fusion	ug/g	95.80	<5.00e-02	3.27e+04	3.18e+04	3.22e+04	2.79	93.60	1.03e+03	n/a
S981003398	F	Lithium - ICP-Fusion	ug/g	92.80	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	92.30	207.0	n/a
S981003398	F	Manganese - ICP-Fusion	ug/g	98.00	<1.00e-01	4.43e+02	451.0	467.0	1.79	95.50	207.0	n/a
S981003398	F	Sodium - ICP-Fusion	ug/g	99.20	<1.00e-01	1.31e+05	1.30e+05	1.31e+05	0.77	86.90	2.07e+03	n/a
S981003398	F	Nickel - ICP-Fusion	ug/g	98.80	7.40e-02	1.94e+04	2.19e+04	2.06e+04	12.1	98.70	4.14e1	n/a
S981003398	F	Silicon - ICP-Fusion	ug/g	107.4	<5.00e-02	7.59e+03	7.93e+03	7.76e+03	4.38	108.0	1.03e+03	n/a
S981003398	F	Uranium - ICP-Fusion	ug/g	98.20	<1.00e-01	3.90e+04	3.83e+04	3.86e+04	1.81	92.50	1.08e+03	n/a
S981003398	F	Zinc - ICP-Fusion	ug/g	96.20	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	97.80	207.0	n/a
S981003398	F	Zirconium - ICP-Fusion	ug/g	101.4	<1.00e-02	<2.06e+02	<2.07e2	n/a	n/a	98.40	207.0	n/a
S981003398	F	Cesium-137 by GEA	uCi/g	97.41	<1.09e-02	2.33e+02	230.0	231.0	1.36	n/a	n/a	1.91
S981003398	F	Alpha of Digested solid	uCi/g	93.87	<1.00e-01	3.44e-01	3.87e-01	3.65e-01	11.8	90.12	2.03e-01	8.24e-01
S981003399	A	1120 by ICP/MS Aci1505159	ug/g	109.0	<1.20e-03	<46.89e+02	<4.67e2	n/a	n/a	n/a	488.5	n/a
S981003399	A	Uranium-234 by ICP/MS Aci1505159	ug/g	n/a	<1.20e-04	<48.86	<4.67e1	n/a	n/a	n/a	48.86	n/a
S981003399	A	Uranium-235 by ICP/MS Aci1505159	ug/g	104.2	<1.20e-04	2.62e+02	302.0	282.1	14.2	94.94	48.86	n/a
S981003399	A	Uranium-236 by ICP/MS Aci1505159	ug/g	n/a	<1.20e-04	<6.55e15	<6.23e3	n/a	n/a	n/a	65.15	n/a
S981003399	A	Uranium-238 by ICP/MS Aci1505159	ug/g	104.0	<1.20e-04	4.19e+02	4.77e+04	4.44e+04	15.5	100.5	48.86	n/a
S981003399	A	Tc99 by ICP/MS Aci1505159	ug/g	93.66	1.47e-04	<8.143	<7.79e0	n/a	n/a	n/a	8.143	n/a

Sample#	R/A#	Unit	Standard %	Blank	Result	Duplicate	Average	RP %	Spk	Rec %	Det	Limit	Count	Err%
S98T003399	A	Sn126 ICP-MS Act/Dig LA-505159	ug/g	<1.50e-03	<0.11e+02	5.8e+2	n/a	n/a	n/a	n/a	610.7	n/a	n/a	n/a
S98T003399	A	Neptunium-237 by ICP-MS	ug/g	119.4	<4.67e-05	<19.01	<1.88e1	n/a	n/a	116.9	19.01	n/a	n/a	n/a
S98T003399	A	Plutonium-239 by ICP-MS	ug/g	102.4	<3.47e-05	<14.13	<1.35e1	n/a	n/a	101.1	14.13	n/a	n/a	n/a
S98T003399	A	Plutonium-240 by ICP-MS	ug/g	n/a	<3.55e-05	<14.47	<1.38e1	n/a	n/a	n/a	14.47	n/a	n/a	n/a
S98T003399	A	Plutonium-242 by ICP-MS	ug/g	n/a	<2.78e-05	<11.32	<1.08e1	n/a	n/a	n/a	11.32	n/a	n/a	n/a
S98T003399	A	Pu/Am-241 by ICP-MS	ug/g	107.8	<3.71e-05	<15.09	<1.44e1	n/a	n/a	106.7	15.09	n/a	n/a	n/a
S98T003399	A	Aluminum -ICP-Acid Digest	ug/g	94.60	1.32e-01	2.03e+04	2.27e+04	2.15e+04	11.2	358.0	20.00	n/a	n/a	n/a
S98T003399	A	Boron -ICP-Acid Digest	ug/g	106.0	7.61e-01	1.17e+02	125.0	121.0	6.61	91.20	20.00	n/a	n/a	n/a
S98T003399	A	Beryllium -ICP-Acid Digest	ug/g	93.60	<5.00e-02	4.04e+02	454.0	429.0	11.7	92.60	20.00	n/a	n/a	n/a
S98T003399	A	Bismuth -ICP-Acid Digest	ug/g	89.80	<1.00e-01	2.86e+03	3.29e+03	3.06e+03	12.8	123.6	40.20	n/a	n/a	n/a
S98T003399	A	Calcium -ICP-Acid Digest	ug/g	97.60	1.65e-01	2.31e+04	2.51e+04	2.41e+04	8.30	332.0	40.20	n/a	n/a	n/a
S98T003399	A	Chromium -ICP-Acid Digest	ug/g	94.00	<1.00e-02	4.79e+02	525.0	502.0	9.16	93.80	4.020	n/a	n/a	n/a
S98T003399	A	Iron -ICP-Acid Digest	ug/g	91.40	<5.00e-02	2.59e+04	2.90e+04	2.74e+04	11.3	400.0	20.00	n/a	n/a	n/a
S98T003399	A	Potassium -ICP-Acid Digest	ug/g	90.60	<5.00e-01	1.11e+03	1.13e+03	1.12e+03	1.79	100.4	20.00	n/a	n/a	n/a
S98T003399	A	Lithium -ICP-Acid Digest	ug/g	95.60	<1.00e-02	<4.010	<3.84e0	n/a	n/a	91.00	4.020	n/a	n/a	n/a
S98T003399	A	Manganese -ICP-Acid Digest	ug/g	92.00	<1.00e-02	3.35e+02	375.0	355.0	11.3	90.80	4.020	n/a	n/a	n/a
S98T003399	A	Sodium -ICP-Acid Digest	ug/g	113.8	1.180	1.19e+05	1.23e+05	1.19e+05	6.72	2.13e+03	40.20	n/a	n/a	n/a
S98T003399	A	Nickel -ICP-Acid Digest	ug/g	91.60	<2.00e-02	1.09e+04	1.21e+04	1.13e+04	4.4	97.0	8.020	n/a	n/a	n/a
S98T003399	A	Silicon -ICP-Acid Digest	ug/g	163.6	2.230	2.86e+03	3.29e+03	3.06e+03	23.1	348.0	20.00	n/a	n/a	n/a
S98T003399	A	Uranium -ICP-Acid Digest	ug/g	92.90	<5.00e-01	3.41e+04	3.68e+04	3.64e+04	10.4	55.40	4.020	n/a	n/a	n/a
S98T003399	A	Zinc -ICP-Acid Digest	ug/g	83.60	1.10e-02	1.18e+02	131.0	124.0	23.9	85.40	4.020	n/a	n/a	n/a
S98T003399	A	Zirconium -ICP-Acid Digest	ug/g	92.40	<1.00e-02	9.200	11.70	10.35	21.9	95.30	1.00e-02	n/a	n/a	n/a
S98T003400	W	Chromium (VI) by Spec.	ug/g	100.8	<27.56	84.70	83.00	76.35	21.9	95.30	1.00e-02	n/a	n/a	n/a
S98T003400	W	OH- by Pot. Titration	ug/g	96.31	<2482.19	4.28e+04	<5.00	247.9	4.84	91.60	134.5	n/a	n/a	n/a
S98T003400	W	Formate by IC-Dioxex 4000/4500	ug/g	95.80	<6.20e-02	2.58e+02	<5.00	247.9	4.84	103.4	49.73	n/a	n/a	n/a
S98T003400	W	Fluoride-IC-Dioxex 4000/4500	ug/g	100.3	<1.20e-02	3.18e2	<1.97e1	1.64e+03	17.9	86.26	70.43	n/a	n/a	n/a
S98T003400	W	Chloride-IC-Dioxex 4000/4500	ug/g	97.86	<1.70e-02	1.89e+02	2.15e+04	2.99e+04	17.4	96.15	447.5	n/a	n/a	n/a
S98T003400	W	Nitrite-IC-Dioxex 4000/4500	ug/g	95.2	<1.25e-01	5.18e+02	<5.18e2	n/a	n/a	85.71	517.9	n/a	n/a	n/a
S98T003400	W	Bromide by Ion Chromatograph	ug/g	106.8	<1.25e-01	3.38e+02	5.39e+04	5.84e+04	18.7	120.8	576.0	n/a	n/a	n/a
S98T003400	W	Nitrate by IC-Dioxex 4000/4500	ug/g	99.6	<1.20e-01	4.13e+02	1.10e+04	1.18e+04	13.6	86.60	497.3	n/a	n/a	n/a
S98T003400	W	Phosphate-IC-Dioxex 4000/4500	ug/g	99.3	<1.20e-01	4.13e+02	2.37e+04	1.99e+04	145	98.75	571.8	n/a	n/a	n/a
S98T003400	W	Sulfate by IC-Dioxex 4000/4500	ug/g	97.33	<1.58e-01	4.38e+03	5.23e+04	3.64e+04	87.8	110.7	435.1	n/a	n/a	n/a
S98T003400	W	Oxalate-IC-Dioxex 4000/4500	ug/g	106.2	<1.00e-01	2.04e+02	5.23e+04	3.64e+04	87.8	110.7	435.1	n/a	n/a	n/a

Table 1. Data Summary Report.
BT-105 (2)CORE NUMBER: 246R
SEGMENT #: 2R

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk. Rec.	Det Limit	Count Err%
S981002806		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	28.20	24.30	26.25	14.9	n/a	n/a
S981002806		DSC Exotherm on Perkin Elmer	Joules/g	100.4	n/a	13.44	11.58	12.51	14.9	n/a	n/a
S981002806		Ammonia by ISE Std Additions	ug/mL	92.71	<5.00	2.76e-02	1.78	2.27	43.2	n/a	55.00
S981002806		Specific Gravity	Sp.G.	98.99	n/a	1.054	1.436	1.445	1.25	n/a	1.00e-03
S981002806		% Water by TOA on Perkin Elmer	%	98.99	n/a	52.73	51.92	52.34	1.60	n/a	n/a
S981002806		Alpha in Liquid Samples	uCi/mL	92.78	<7.88e-03	4.29e-03	<6.91E-3	n/a	n/a	1.00e-02	5.00E-02
S981002814		Strontium-89/90 High Level	uCi/mL	99.04	<2.01e-04	1.30e-01	1.30e-01	1.30e-01	0.00	n/a	1.80E-00
S981002814		Pu-239/240 by TRU-SPEC Resin	uCi/mL	112.5	<3.19e-05	3.67e-02	<3.58E-5	n/a	n/a	2.07e-04	1.22E-01
S981002814		Aluminum-ICP-Acid Dil.	ug/mL	97.20	<3.05e-02	2.07e-02	2.03e+04	2.02e+04	0.99	n/a	30.10
S981002814		Boron-ICP-Acid Dil.	ug/mL	102.8	<3.05e-02	32.40	31.80	32.10	n/a	n/a	30.10
S981002814		Barium-ICP-Acid Dil.	ug/mL	96.80	<3.05e-02	<6.10	<3.01e1	n/a	n/a	n/a	30.10
S981002814		Bismuth-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10
S981002814		Calcium-ICP-Acid Dil.	ug/mL	103.0	<1.00e-01	1.52e+03	1.51e+03	1.51e+03	1.32	n/a	6.010
S981002814		Chromium-ICP-Acid Dil.	ug/mL	99.40	<1.00e-02	3.0e-03	<3.01e3	4.02e+03	1.24	n/a	300.0
S981002814		Iron-ICP-Acid Dil.	ug/mL	95.60	<3.05e-02	4.00e+03	4.05e+03	4.02e+03	n/a	n/a	6.010
S981002814		Potassium-ICP-Acid Dil.	ug/mL	100.6	<1.00e-01	<6.10	<6.01e0	n/a	n/a	n/a	6.010
S981002814		Lithium-ICP-Acid Dil.	ug/mL	99.60	<1.00e-02	<6.10	<6.01e0	n/a	n/a	n/a	6.010
S981002814		Manganese-ICP-Acid Dil.	ug/mL	98.40	<1.00e-01	2.01e+05	2.00e+05	2.01e+05	0.50	n/a	12.00
S981002814		Sodium-ICP-Acid Dil.	ug/mL	98.80	<2.00e-02	1.86e-01	<1.20e1	193.5	7.75	n/a	30.10
S981002814		Nickel-ICP-Acid Dil.	ug/mL	103.2	<3.05e-02	2.01e+02	<3.01e2	n/a	n/a	n/a	300.0
S981002814		Silicon-ICP-Acid Dil.	ug/mL	96.90	<3.05e-02	<6.10	<6.01e0	n/a	n/a	n/a	6.010
S981002814		Uranium-ICP-Acid Dil.	ug/mL	94.20	<1.00e-02	<6.10	<6.01e0	n/a	n/a	n/a	6.010
S981002814		Zinc-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	<6.10	<6.01e0	n/a	n/a	n/a	6.010
S981002814		Zirconium-ICP-Acid Dil.	ug/mL	107.6	<1.00e-02	3.04e+03	3.05e+03	3.04e+03	0.33	40.25	68.88
S981002814		Formate by Ion Chromatograph	ug/mL	104.4	<1.20e-02	3.18e+02	1.07	507.6	4.14	n/a	n/a
S981002814		Fluoride-IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	3.99e+03	3.99e+03	3.99e+03	1.00	n/a	87.57
S981002814		Chloride-IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	3.99e+03	3.99e+03	3.99e+03	1.00	n/a	556.3
S981002814		Nitrate-IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	3.99e+03	3.99e+03	3.99e+03	1.00	n/a	643.9
S981002814		Bromide by Ion Chromatograph	ug/mL	99.52	<1.25e-01	1.12e+03	1.14e+03	1.14e+03	1.58	n/a	716.0
S981002814		Nitrate by IC-Dionex 4000/4500	ug/mL	98.85	<1.99e-01	2.35e+03	2.35e+03	2.35e+03	1.58	n/a	618.1
S981002814		Phosphate by IC-Dionex 4000/4500	ug/mL	99.53	<1.20e-01	1.05e+03	1.12e+03	1.08e+03	1.39	n/a	710.8
S981002814		Sulfate by IC-Dionex 4000/4500	ug/mL	100.4	<1.38e-01	2.14e+03	2.17e+03	2.16e+03	1.39	n/a	540.9
S981002814		Oxalate-IC-Dionex 4000/4500	ug/mL	102.2	<1.00e-01	<5.41e+02	<5.41e2	n/a	n/a	n/a	0.140
S981002814		Cesium-137 by GEA	uCi/mL	103.4	<1.57e-03	1.62e+02	1.62	162.0	0.00	n/a	1.00e-02
S981002814		Chromium (VI) by Spec.	ug/mL	102.0	<2.58e-01	13.40	13.50	13.45	0.74	n/a	4.860
S981002822		1.129 by ICP/MS Acid Addition	ug/mL	104.7	<1.20e-04	<4.860	<4.860	n/a	n/a	n/a	62.50
S981002822		08- by Pot. Titration	ug/mL	93.33	<3.200	2.13e+03	2.16e+03	2.16e+03	3.23	n/a	2.053
S981002822		TOC by Persulfate/Coulometry	ug/mL	101.7	<5.07e-05	<2.053	<2.050	n/a	n/a	n/a	1.659
S981002822		Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<4.10e-05	<1.659	<1.660	n/a	n/a	n/a	1.845
S981002822		Uranium-235 by ICP/MS Acid Add	ug/mL	102.1	<4.55e-05	<1.645	<1.448	n/a	n/a	n/a	2.373
S981002822		Uranium-236 by ICP/MS Acid Add	ug/mL	105.0	<3.57e-05	<2.373	<2.376	n/a	n/a	n/a	8.10e-02
S981002822		Uranium-238 by ICP/MS Acid Add	ug/mL	81.73	<2.00e-06	4.001	4.120	4.061	2.96	n/a	n/a
S981002822		P 99 by ICP/MS Acid Addition	ug/mL								

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98T002822	D	Sn 126	By ICP/MS	ug/mL										
S98T002822	A	Neptunium-237	By ICP/MS	ug/mL	134.1	<1.50e-04	< 6.07e-01	<5.06e-01	N/A	N/A	N/A	6.53e-01	N/A	N/A
S98T002822	A	Neptunium-237	By ICP/MS	ug/mL	134.1	N/A	< 6.53e-01	<5.53e-01	N/A	N/A	N/A	6.53e-01	N/A	N/A
S98T002822	A	Plutonium-239	By ICP/MS	ug/mL	137.0	<1.75e-05	< 7.07e-01	<7.09e-01	N/A	N/A	N/A	7.09e-01	N/A	N/A
S98T002822	A	Plutonium-240	By ICP/MS	ug/mL	137.0	<1.02e-05	< 7.14e-01	<7.14e-01	N/A	N/A	N/A	7.14e-01	N/A	N/A
S98T002822	A	Plutonium-242	By ICP/MS	ug/mL	137.0	<1.39e-05	< 9.48e-01	<9.48e-01	N/A	N/A	N/A	9.48e-01	N/A	N/A
S98T002822	A	Plutonium-241	By ICP/MS	ug/mL	137.0	<1.58e-05	< 6.41e-01	<6.41e-01	N/A	N/A	N/A	6.41e-01	N/A	N/A

Table 1. Data Summary Report.
BT-105 (2)

CORE NUMBER: 246R
SEGMENT #: 3R

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002807		DSC Exotherm DRY Calculated	Joules/g Dry	n/a	n/a	23.51	45.72	26.55	4.7	n/a	n/a
S98T002807		DSC Exotherm on Perkin Elmer	Joules/g	97.15	n/a	11.75	12.52	12.23	4.2	n/a	n/a
S98T002807		Ammonia by ISE-STD Additions	ug/mL	92.71	<5.00	1.86e+02	61.20	155.3	4.02	55.00	n/a
S98T002807		Specific Gravity	Sp/g	98.97	n/a	1.572	93.32	1.432	1.12	1.00e-03	n/a
S98T002807		% Water by IGA on Perkin Elmer	%	98.94	n/a	55.35	51.30	53.19	0.56	n/a	n/a
S98T002807		Alpha in Liquid Samples	uCi/mL	92.72	<7.58e-02	<1.03e-02	<1.38E-2	<1.38E-2	2.74	2.00e-02	5.00E+02
S98T002807		Strontium-89/90 High Level	uCi/mL	92.72	<1.42e-01	<1.42e-01	<1.42e-01	<1.42e-01	2.74	2.00e-02	1.71E+02
S98T002815		Pd-239/240 by TMO-SPEC Resin	uCi/mL	112.5	<19.0e-04	<4.22e-05	<3.87E-5	<3.87E-5	n/a	4.22e-05	1.00E+02
S98T002815		D Aluminum-ICP-Acid DIL	ug/mL	97.30	<5.00e-02	3.06e+02	3.06e+02	3.05e+04	0.65	n/a	n/a
S98T002815		D Boron-ICP-Acid DIL	ug/mL	102.8	<5.00e-02	<30.30	<35.80	37.55	9.32	n/a	n/a
S98T002815		D Barium-ICP-Acid DIL	ug/mL	98.90	<5.00e-02	<60.10	<60.10	n/a	n/a	30.10	n/a
S98T002815		D Bismuth-ICP-Acid DIL	ug/mL	100.4	<1.00e-01	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002815		D Calcium-ICP-Acid DIL	ug/mL	103.0	<1.00e-01	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002815		D Chlorine-ICP-Acid DIL	ug/mL	99.40	<5.00e-02	<30.10	<33.01e1	1.54e+03	0.65	n/a	n/a
S98T002815		D Cobalt-ICP-Acid DIL	ug/mL	95.60	<5.00e-02	<18.0e-03	<18.0e-03	4.16e+03	0.96	n/a	n/a
S98T002815		D Lithium-ICP-Acid DIL	ug/mL	99.60	<1.00e-01	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002815		D Manganese-ICP-Acid DIL	ug/mL	98.40	<1.00e-02	<60.10	<60.10	1.98e+05	1.52	n/a	n/a
S98T002815		D Nickel-ICP-Acid DIL	ug/mL	98.80	<2.00e-02	<12.00	<11.20e1	n/a	n/a	12.00	n/a
S98T002815		D Silicon-ICP-Acid DIL	ug/mL	103.2	<5.00e-02	<2.43e+02	<3.01e2	231.0	10.4	n/a	n/a
S98T002815		D Uranium-ICP-Acid DIL	ug/mL	96.90	<5.00e-02	<60.10	<60.10	n/a	n/a	300.0	n/a
S98T002815		D Zinc-ICP-Acid DIL	ug/mL	94.20	<1.00e-02	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002815		D Zirconium-ICP-Acid DIL	ug/mL	97.20	<1.00e-02	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002815		Formate by Ion Chromatograph	ug/mL	96.64	<4.50e-02	<1.76e+03	<1.82e+03	1.79e+03	3.35	n/a	37.57
S98T002815		Fluoride-IC-Dionex 4000/4500	ug/mL	104.4	<1.20e-02	<4.01e+02	<377.0	395.2	6.17	n/a	n/a
S98T002815		Chloride-IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	<5.45e+03	<5.92e+03	5.69e+03	8.27	n/a	87.57
S98T002815		Nitrate by Ion Chromatograph	ug/mL	102.5	<1.08e-01	<6.22e+04	<6.42e+04	6.32e+04	3.16	n/a	556.3
S98T002815		Bromide by Ion Chromatograph	ug/mL	99.32	<1.29e-01	<8.41e+02	<868.0	854.4	3.16	n/a	643.9
S98T002815		Nitrate by IC-Dionex 4000/4500	ug/mL	98.85	<1.39e-01	<2.25e+05	<2.25e+05	2.24e+05	0.89	n/a	716.0
S98T002815		Phosphate by IC-Dionex 4000/4500	ug/mL	99.53	<1.20e-01	<8.27e+02	<828.0	827.4	0.12	n/a	618.1
S98T002815		Sulfate by IC-Dionex 4000/4500	ug/mL	100.4	<1.38e-01	<2.18e+03	<2.26e+03	2.22e+03	3.60	n/a	710.8
S98T002815		Oxalate by IC-Dionex 4000/4500	ug/mL	102.2	<1.03e-01	<5.41e+02	<5.41e2	n/a	n/a	540.9	n/a
S98T002815		Chromium (VI) by Spec.	ug/mL	103.4	<1.57e-05	<1.77e+02	<177.0	177.0	0.00	n/a	0.140
S98T002823		D 1. 129 by ICP/MS Acid Addition	ug/mL	103.0	<2.58e-01	<14.40	<14.50	14.45	0.69	n/a	1.00e-02
S98T002823		D OH- by Pot. Titration	ug/mL	104.7	<1.20e-04	<4.86e0	<4.86e0	n/a	n/a	4.860	n/a
S98T002823		TOC by Persulfate/Coulometry	ug/mL	101.7	<3.200	<2.31e+03	<2.03e+03	1.02e+03	7.33	n/a	62.50
S98T002823		D Uranium-235 by ICP/MS Acid Add	ug/mL	93.33	<5.07e-05	<2.053	<2.05e0	2.17e-03	12.9	n/a	40.00
S98T002823		D Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<4.10e-05	<1.659	<1.66e0	n/a	n/a	2.053	n/a
S98T002823		D Uranium-235 by ICP/MS Acid Add	ug/mL	102.1	<4.55e-05	<1.845	<1.84e0	n/a	n/a	1.845	n/a
S98T002823		D Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<5.57e-05	<1.448	<1.45e0	n/a	n/a	1.448	n/a
S98T002823		D Uranium-238 by ICP/MS Acid Add	ug/mL	105.0	<5.86e-05	<2.373	<2.37e0	n/a	n/a	2.373	n/a
S98T002823		D Tc 99 by ICP/MS Acid Addition	ug/mL	81.73	<2.00e-06	<4.356	<4.430	4.408	0.91	n/a	8.10e-02

Sample#	R/A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002823	D	Sn 126 by ICP/MS Acid Addition	ug/mL	n/a	<1.50e-04	< 6.075	<6.08e0	n/a	n/a	n/a	6.075	n/a
S98T002823	A	Neptunium-237 by ICP/MS	ug/mL	134.1	n/a	<6.53e-01	<6.53e-1	n/a	n/a	n/a	6.53e-01	n/a
S98T002823	A	Plutonium-239 by ICP/MS	ug/mL	139.0	<1.75e-05	<7.09e-01	<7.09e-1	n/a	n/a	n/a	7.09e-01	n/a
S98T002823	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<1.02e-05	<4.14e-01	<4.14e-1	n/a	n/a	n/a	4.14e-01	n/a
S98T002823	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<2.39e-05	<9.68e-01	<9.68e-1	n/a	n/a	n/a	9.68e-01	n/a
S98T002823	A	Pu/Am-241 by ICP/MS	ug/mL	n/a	<1.58e-05	<6.41e-01	<6.41e-1	n/a	n/a	n/a	6.41e-01	n/a

Table 1. Data Summary Report.
BT-05 (2)

CORE NUMBER: 246R
SEGMENT #: 4R

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98T002808		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	20.48	23.08	21.78	11.9	n/a	n/a		n/a
S98T002808		DSC Exotherm on Perkin Elmer	Joules/g	97.15	n/a	9.658	10.88	10.27	11.9	n/a	10.00		n/a
S98T002808		Ammonia by ISE-Std Additions	SP-g.	97.55	<10.00	1.17e+02	102.0	114.0	0.35	n/a	1.00e+03		n/a
S98T002808		Specific Gravity	SP-g.	98.97	n/a	1.444	1.437	1.442	0.35	n/a	1.00e+03		n/a
S98T002808		% Water by TGA on Perkin Elmer	SP-g.	98.97	n/a	32.86	31.54	32.25	0.35	n/a	1.00e+03		n/a
S98T002808		Alpha in Liquid Samples	UCI/mL	92.78	<7.58e-05	1.35e+02	1.25e+02	1.35e+02	0.35	n/a	2.00e+02		2.34e+02
S98T002808		Strontium-89/90 High Level	UCI/mL	97.04	<2.01e-04	1.69e+01	1.69e+01	1.69e+01	3.08	n/a	2.07e+04		1.40E+00
S98T002816		Pu-239/240 by TRU-SPEC Resin	UCI/mL	126.3	<3.19e-03	3.75e+02	4.45e+02	4.10e+02	1.9	n/a	3.75e+05		1.00E+02
S98T002816		D Aluminum-IP-Acid DIL.	ug/mL	97.80	<5.00e-02	3.35e+04	3.35e+04	3.61e+04	3.43	n/a	30.10		n/a
S98T002816		D Barium-IP-Acid DIL.	ug/mL	98.40	<5.00e-02	32.60	31.50	32.05	3.43	n/a	30.10		n/a
S98T002816		D Barium-IP-Acid DIL.	ug/mL	98.40	<5.00e-02	32.60	31.50	32.05	3.43	n/a	30.10		n/a
S98T002816		D Barium-IP-Acid DIL.	ug/mL	98.40	<5.00e-02	32.60	31.50	32.05	3.43	n/a	30.10		n/a
S98T002816		D Calcium-IP-Acid DIL.	ug/mL	100.0	<1.00e-01	60.10	<6.01e1	n/a	n/a	n/a	60.10		n/a
S98T002816		D Calcium-IP-Acid DIL.	ug/mL	100.0	<1.00e-01	60.10	<6.01e1	n/a	n/a	n/a	60.10		n/a
S98T002816		D Chromium-IP-Acid DIL.	ug/mL	103.2	<1.00e-02	1.64e+03	1.64e+03	1.64e+03	0.51	n/a	6.010		n/a
S98T002816		D Chromium-IP-Acid DIL.	ug/mL	103.2	<1.00e-02	1.64e+03	1.64e+03	1.64e+03	0.51	n/a	6.010		n/a
S98T002816		D Iron-IP-Acid DIL.	ug/mL	96.40	<5.00e-01	4.42e+03	4.48e+03	4.45e+03	1.35	n/a	30.10		n/a
S98T002816		D Potassium-IP-Acid DIL.	ug/mL	174.0	<5.00e-01	<5.010	<5.010	n/a	n/a	n/a	6.010		n/a
S98T002816		D Lithium-IP-Acid DIL.	ug/mL	98.40	<1.00e-02	<5.010	<5.010	n/a	n/a	n/a	6.010		n/a
S98T002816		D Manganese-IP-Acid DIL.	ug/mL	98.40	<1.00e-02	2.04e+05	2.04e+05	2.05e+05	0.98	n/a	60.10		n/a
S98T002816		D Sodium-IP-Acid DIL.	ug/mL	98.40	<1.00e-02	<12.00	<12.00	n/a	n/a	n/a	12.00		n/a
S98T002816		D Nickel-IP-Acid DIL.	ug/mL	106.4	<5.00e-02	2.39e+02	2.57e+02	2.48e+02	7.26	n/a	30.10		n/a
S98T002816		D Silicon-IP-Acid DIL.	ug/mL	96.10	<5.00e-02	<5.01e+02	<5.01e+02	n/a	n/a	n/a	300.0		n/a
S98T002816		D Uranium-IP-Acid DIL.	ug/mL	92.80	<1.00e-02	<5.010	<5.010	n/a	n/a	n/a	6.010		n/a
S98T002816		D Zinc-IP-Acid DIL.	ug/mL	95.60	<1.00e-02	<5.010	<5.010	n/a	n/a	n/a	6.010		n/a
S98T002816		D Zirconium-IP-Acid DIL.	ug/mL	92.31	<2.10e-02	1.04e+03	1.04e+03	1.05e+03	1.90	n/a	319.4		n/a
S98T002816		Formate by Ion Chromatograph	ug/mL	106.4	<1.20e-02	2.94e+02	2.97e+02	2.95e+02	1.02	n/a	61.81		n/a
S98T002816		Fluoride by IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	4.69e+03	4.69e+03	4.69e+03	0.00	n/a	87.57		n/a
S98T002816		Chloride by IC-Dionex 4000/4500	ug/mL	102.0	<1.70e-02	4.69e+03	4.69e+03	4.69e+03	0.00	n/a	87.57		n/a
S98T002816		Nitrate by IC-Dionex 4000/4500	ug/mL	102.5	<1.08e-01	6.44e+04	6.40e+04	6.42e+04	0.60	n/a	556.3		n/a
S98T002816		Bromide by Ion Chromatograph	ug/mL	99.52	<1.29e-01	7.15e+02	6.94e+02	7.04e+02	2.98	n/a	643.9		n/a
S98T002816		Nitrate by IC-Dionex 4000/4500	ug/mL	98.85	<1.39e-01	2.09e+05	2.09e+05	2.09e+05	0.00	n/a	716.0		n/a
S98T002816		Phosphate by IC-Dionex 4000/4500	ug/mL	99.53	<1.39e-01	<6.18e+02	<6.18e+02	n/a	n/a	n/a	618.1		n/a
S98T002816		Sulfate by IC-Dionex 4000/4500	ug/mL	100.4	<1.38e-01	2.12e+03	2.07e+03	2.09e+03	2.39	n/a	710.8		n/a
S98T002816		Oxalate by Dionex 4000/450	ug/mL	103.4	<1.57e-05	<5.11e+02	<5.11e+02	n/a	n/a	n/a	540.9		n/a
S98T002816		Cesium-137 by GEA	UCI/mL	100.4	<2.58e-01	1.40e+03	1.40e+03	1.40e+03	0.00	n/a	1.00e+02		0.130
S98T002824		Chromium (VI) by Spec.	ug/mL	106.7	<1.20e-04	<4.840	<4.840	n/a	n/a	n/a	6.250		n/a
S98T002824		D 1, 1, 2 by IC/MS Acid Addition	ug/mL	101.7	<1.20e-04	<4.840	<4.840	n/a	n/a	n/a	6.250		n/a
S98T002824		OH- by Pot. Titration	ug/mL	93.33	<1.20e-05	2.41e+03	2.40e+03	2.40e+03	0.42	n/a	4.86e+01		n/a
S98T002824		DOC by Persulfate/Oxidometry	ug/mL	n/a	<1.20e-05	<4.86e+01	<4.86e+01	n/a	n/a	n/a	4.86e+01		n/a
S98T002824		D Uranium-235 by IC/MS Acid Add	ug/mL	100.6	<1.20e-05	<4.86e+01	<4.86e+01	n/a	n/a	n/a	4.86e+01		n/a
S98T002824		D Uranium-235 by IC/MS Acid Add	ug/mL	100.6	<1.20e-05	<4.86e+01	<4.86e+01	n/a	n/a	n/a	4.86e+01		n/a
S98T002824		D Uranium-235 by IC/MS Acid Add	ug/mL	101.0	<1.20e-05	<4.86e+01	<4.86e+01	n/a	n/a	n/a	4.86e+01		n/a
S98T002824		D Uranium-235 by IC/MS Acid Add	ug/mL	101.0	<1.20e-05	<4.86e+01	<4.86e+01	n/a	n/a	n/a	4.86e+01		n/a
S98T002824		D Uranium-238 by IC/MS Acid Add	ug/mL	101.0	<1.20e-05	7.27e+01	6.84e+01	7.27e+01	11.7	n/a	8.10e+02		n/a
S98T002824		D Uranium-238 by IC/MS Acid Add	ug/mL	78.68	<2.00e-06	4.880	5.050	4.965	3.42	n/a	8.10e+02		n/a
S98T002824		D TC by IC/MS Acid Addition	ug/mL										

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPO %	Spk Rec %	Det Limit	Count Err%
S98T002824	D	Sn 128 By ICP/MS	Acid Addition	ug/mL								
S98T002824	A	Neptunium-237 By ICP/MS	ug/mL	n/a	<1.50e-04	< 6.075	<6.08e0	n/a	n/a	n/a	6.075	n/a
S98T002824	A	Neptunium-237 By ICP/MS	ug/mL	134.1	n/a	<6.53e-01	<6.53e-1	n/a	n/a	n/a	6.53e-01	n/a
S98T002824	A	Plutonium-239 By ICP/MS	ug/mL	139.0	<1.75e-05	<7.09e-01	<7.09e-1	n/a	n/a	n/a	7.09e-01	n/a
S98T002824	A	Plutonium-240 By ICP/MS	ug/mL	n/a	<1.02e-05	<4.18e-01	<4.18e-1	n/a	n/a	n/a	4.18e-01	n/a
S98T002824	A	Plutonium-242 By ICP/MS	ug/mL	n/a	<2.39e-05	<9.68e-01	<9.68e-1	n/a	n/a	n/a	9.68e-01	n/a
S98T002824	A	Pu/Am-241 By ICP/MS	ug/mL	n/a	<1.58e-05	<6.41e-01	<6.41e-1	n/a	n/a	n/a	6.41e-01	n/a

Table 1. Data Summary Report.
81-105 (2)CORE NUMBER: 246R
SEGMENT #: 5R

SEGMENT POSITION: Drainable Liquid		Unit		Standard %	Blank	Result	Duplicate	Average	RPD % Spk	%	Det Limit	Count Err%
Sample#	R# Analyte	Joules/g Dry	Joules/g									
S981002809	DSC Exotherm Dry Calculated	Joules/g	34.36	58.04	51.3	58.04	51.3	58.04	51.3	n/a	n/a	n/a
S981002809	DSC Exotherm Dry Calculated	Joules/g	34.36	58.04	51.3	58.04	51.3	58.04	51.3	n/a	n/a	n/a
S981002809	DSC Exotherm Dry Calculated	Joules/g	34.36	58.04	51.3	58.04	51.3	58.04	51.3	n/a	n/a	n/a
S981002809	D I 129 by ICP/MS Acid Addition	ug/mL	4.860	< 4.860	< 4.860	< 4.860	< 4.860	< 4.860	< 4.860	n/a	4.860	n/a
S981002809	Ammonia by ISE-Std Additions	ug/mL	1.91e-02	188.0	1.91e-02	188.0	1.91e-02	188.0	1.91e-02	n/a	1.91e-02	n/a
S981002809	Specific Gravity	SP-G.	1.445	1.445	1.445	1.445	1.445	1.445	1.445	n/a	1.00e-03	n/a
S981002809	% Water by IGA on Perkin Elmer	%	53.99	53.99	53.99	53.99	53.99	53.99	53.99	n/a	2.00e-02	8.58e-01
S981002809	Alpha in Liquid Samples	uCi/mL	92.78	< 7.58e-03	1.77e-02	1.77e-02	1.77e-02	1.77e-02	1.77e-02	n/a	2.00e-02	n/a
S981002809	I 129 by ICP/MS Acid Addition	ug/mL	4.860	< 4.860	< 4.860	< 4.860	< 4.860	< 4.860	< 4.860	n/a	4.860	n/a
S981002817	Strontium-89/90 High Level	uCi/mL	99.04	< 2.01e-04	1.57e-01	1.58e-01	1.58e-01	1.58e-01	1.58e-01	n/a	2.04e-04	1.63e+00
S981002817	Pu-239/240 by TRU-SPEC Resin	uCi/mL	112.5	< 3.19e-06	4.03e-05	4.14E-5	4.14E-5	4.14E-5	4.14E-5	n/a	4.03e-05	1.22e+01
S981002817	Aluminum-ICP-Acid Dtl.	ug/mL	97.80	< 5.00e-02	3.26e+04	3.26e+04	3.26e+04	3.26e+04	3.26e+04	n/a	3.01e-01	n/a
S981002817	Boron-ICP-Acid Dtl.	ug/mL	102.6	< 5.00e-02	32.60	33.80	33.80	33.80	33.80	n/a	30.10	n/a
S981002817	Barium-ICP-Acid Dtl.	ug/mL	96.40	< 5.00e-02	< 30.10	< 30.10	< 30.10	< 30.10	< 30.10	n/a	30.10	n/a
S981002817	Bismuth-ICP-Acid Dtl.	ug/mL	100.0	< 1.00e-01	< 60.10	< 60.10	< 60.10	< 60.10	< 60.10	n/a	60.10	n/a
S981002817	Calcium-ICP-Acid Dtl.	ug/mL	103.2	< 1.00e-01	< 60.10	< 60.10	< 60.10	< 60.10	< 60.10	n/a	60.10	n/a
S981002817	Chromium-ICP-Acid Dtl.	ug/mL	96.40	< 5.00e-02	< 30.10	< 30.10	< 30.10	< 30.10	< 30.10	n/a	30.10	n/a
S981002817	Iron-ICP-Acid Dtl.	ug/mL	97.20	< 5.00e-01	4.69e+03	4.69e+03	4.69e+03	4.69e+03	4.69e+03	n/a	300.0	n/a
S981002817	Potassium-ICP-Acid Dtl.	ug/mL	100.4	< 1.00e-02	< 6.010	< 6.010	< 6.010	< 6.010	< 6.010	n/a	6.010	n/a
S981002817	Lithium-ICP-Acid Dtl.	ug/mL	98.40	< 1.00e-02	< 6.010	< 6.010	< 6.010	< 6.010	< 6.010	n/a	6.010	n/a
S981002817	Manganese-ICP-Acid Dtl.	ug/mL	98.40	< 1.00e-02	< 6.010	< 6.010	< 6.010	< 6.010	< 6.010	n/a	6.010	n/a
S981002817	Sodium-ICP-Acid Dtl.	ug/mL	98.20	< 1.00e-02	< 12.00	< 12.00	< 12.00	< 12.00	< 12.00	n/a	12.00	n/a
S981002817	Nickel-ICP-Acid Dtl.	ug/mL	104.4	< 5.00e-02	< 3.01e-02	< 3.01e-02	< 3.01e-02	< 3.01e-02	< 3.01e-02	n/a	30.10	n/a
S981002817	Silicon-ICP-Acid Dtl.	ug/mL	96.10	< 5.00e-01	2.11e+05	2.11e+05	2.11e+05	2.11e+05	2.11e+05	n/a	60.10	n/a
S981002817	Uranium-ICP-Acid Dtl.	ug/mL	94.80	< 1.00e-02	< 6.010	< 6.010	< 6.010	< 6.010	< 6.010	n/a	6.010	n/a
S981002817	Zinc-ICP-Acid Dtl.	ug/mL	94.80	< 1.00e-02	< 6.010	< 6.010	< 6.010	< 6.010	< 6.010	n/a	6.010	n/a
S981002817	Zirconium-ICP-Acid Dtl.	ug/mL	92.31	< 2.10e-02	1.13e+03	1.13e+03	1.13e+03	1.13e+03	1.13e+03	n/a	319.4	n/a
S981002817	Formate by Ion Chromatograph	ug/mL	103.1	< 1.20e-02	395.0	395.0	395.0	395.0	395.0	n/a	61.81	n/a
S981002817	Fluoride-IC-Dioxex 4000/4500	ug/mL	100.9	< 1.70e-02	4.56e+03	4.56e+03	4.56e+03	4.56e+03	4.56e+03	n/a	87.57	n/a
S981002817	Chloride-IC-Dioxex 4000/4500	ug/mL	104.1	< 1.08e-01	6.48e+04	6.48e+04	6.48e+04	6.48e+04	6.48e+04	n/a	556.3	n/a
S981002817	Nitrate-IC by Ion Chromatograph	ug/mL	98.94	< 1.25e-01	6.47e+02	654.0	654.0	654.0	654.0	n/a	643.9	n/a
S981002817	Bromide by IC-Dioxex 4000/4500	ug/mL	98.18	< 1.39e-01	2.14e+05	2.14e+05	2.14e+05	2.14e+05	2.14e+05	n/a	716.0	n/a
S981002817	Nitrate by IC-Dioxex 4000/4500	ug/mL	100.5	< 1.20e-01	1.77e+02	1.77e+02	1.77e+02	1.77e+02	1.77e+02	n/a	618.1	n/a
S981002817	Phosphate-IC-Dioxex 4000/4500	ug/mL	101.4	< 1.38e-01	9.79e+03	9.79e+03	9.79e+03	9.79e+03	9.79e+03	n/a	710.8	n/a
S981002817	Sulfate by IC-Dioxex 4000/4500	ug/mL	106.6	< 1.57e-01	5.41e+02	5.41e+02	5.41e+02	5.41e+02	5.41e+02	n/a	540.9	n/a
S981002817	Oxalate by IC-Dioxex 4000/4500	ug/mL	103.4	< 1.57e-01	1.90e+02	191.0	191.0	191.0	191.0	n/a	0.130	n/a
S981002817	Cesium-137 by GEA	ug/mL	100.4	< 2.58e-01	1.51e+03	1.51e+03	1.51e+03	1.51e+03	1.51e+03	n/a	1.00e-02	n/a
S981002817	Chromium (VI) by Spec.	ug/mL	101.7	< 1.03e-03	1.03e+03	1.03e+03	1.03e+03	1.03e+03	1.03e+03	n/a	62.50	n/a
S981002817	OH- by Pot. Titration	ug/mL	92.33	< 6.100	2.32e+03	2.32e+03	2.32e+03	2.32e+03	2.32e+03	n/a	40.00	n/a
S981002825	TOC by Persulfate/Oxidometry	ug/mL	n/a	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a
S981002825	Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a
S981002825	Uranium-234 by ICP/MS Acid Add	ug/mL	100.6	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a
S981002825	Uranium-235 by ICP/MS Acid Add	ug/mL	n/a	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a
S981002825	Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a
S981002825	Uranium-238 by ICP/MS Acid Add	ug/mL	101.0	< 1.20e-05	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	< 4.86e-01	n/a	4.86e-01	n/a

Sample#	P	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002825		D Te 99 by ICP/MS	ug/mL	78.68	<2.00e-06	4.560	4.460	4.510	2.22	N/A	8.10e-02	N/A
S981002825		D Sn 126 by ICP/MS	ug/mL	N/A	<1.50e-04	< 6.075	<6.0860	N/A	N/A	N/A	6.075	N/A
S981002825		A Rextunium-237 by ICP/MS	ug/mL	134.1	N/A	<6.53e-01	<6.53e-01	N/A	N/A	N/A	6.53e-01	N/A
S981002825		A Plutonium-239 by ICP/MS	ug/mL	139.0	<1.75e-05	<7.09e-01	<7.09e-01	N/A	N/A	N/A	7.09e-01	N/A
S981002825		A Plutonium-240 by ICP/MS	ug/mL	N/A	<1.02e-05	<4.14e-01	<4.14e-01	N/A	N/A	N/A	4.14e-01	N/A
S981002825		A Plutonium-242 by ICP/MS	ug/mL	N/A	<2.39e-05	<9.68e-01	<9.68e-01	N/A	N/A	N/A	9.68e-01	N/A
S981002825		A Pu/Am-241 by ICP/MS	ug/mL	N/A	<1.58e-05	<6.41e-01	<6.41e-01	N/A	N/A	N/A	6.41e-01	N/A

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 246R
SEGMENT #: 6R

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Def Limit	Count Err%
S98T002810	D	DSC Exotherm Dry	Joules/g Dry	n/a	n/a	46.67	53.46	50.06	13.6	n/a	n/a	n/a
S98T002810	D	DSC Exotherm on Perkin Elmer	Joules/g	95.36	n/a	22.05	25.26	23.66	13.6	n/a	n/a	n/a
S98T002810	D	Ammonia by ISE-std Additions	Sp/g	97.00	<25.00	1.07e+02	137.0	122.0	24.6	n/a	25.00	n/a
S98T002810	D	Specific Gravity	Sp/g	99.06	n/a	1.999	1.411	1.405	0.85	n/a	1.00e-03	n/a
S98T002810	D	% Water by TGA on Perkin Elmer	%	98.64	n/a	52.72	32.78	52.75	0.11	n/a	n/a	n/a
S98T002810	D	Alpha In Liquid Samples	UCI/mL	92.78	<7.58e-03	<1.58e-02	<2.17E-2	<1.58e-02	n/a	n/a	2.00e-02	2.34e+02
S98T002810	D	Strontium-89/90 High Level	UCI/mL	101.9	<2.00e-03	1.68e-01	1.88e-01	1.68e-01	0.00	n/a	2.00e-03	4.67e+00
S98T002810	D	Pu-239/240 by TRU-SPEC Resin	UCI/mL	105.4	<3.35e-04	<3.47E-4	<3.47E-4	<3.47E-4	n/a	n/a	3.49e-04	1.04e+01
S98T002818	D	Aluminum-ICP-Acid DIL	ug/mL	97.60	<5.00e-02	4.49e+04	4.50e+04	4.50e+04	0.22	n/a	30.10	n/a
S98T002818	D	Boron-ICP-Acid DIL	ug/mL	101.2	<5.00e-02	<30.10	31.20	<30.10	n/a	n/a	30.10	n/a
S98T002818	D	Barium-ICP-Acid DIL	ug/mL	98.60	<5.00e-02	<30.10	<30.10	<30.10	n/a	n/a	30.10	n/a
S98T002818	D	Bismuth-ICP-Acid DIL	ug/mL	100.2	<1.00e-01	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Calcium-ICP-Acid DIL	ug/mL	103.2	<1.00e-01	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Chromium-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	1.69e+03	1.68e+03	1.69e+03	0.01	n/a	30.10	n/a
S98T002818	D	Copper-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	30.30	30.70	30.30	0.31	n/a	30.10	n/a
S98T002818	D	Fluorine-ICP-Acid DIL	ug/mL	99.60	<1.00e-01	4.99e+03	4.92e+03	4.96e+03	0.40	n/a	30.10	n/a
S98T002818	D	Lithium-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Magnesium-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Manganese-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Nickel-ICP-Acid DIL	ug/mL	99.60	<1.00e-02	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Silicon-ICP-Acid DIL	ug/mL	103.8	<5.00e-02	<2.64e+02	262.0	<2.64e+02	0.46	n/a	30.10	n/a
S98T002818	D	Zinc-ICP-Acid DIL	ug/mL	96.20	<1.00e-01	<30.10	<30.10	<30.10	0.76	n/a	30.10	n/a
S98T002818	D	Zirconium-ICP-Acid DIL	ug/mL	96.80	<1.00e-02	<60.10	<60.10	<60.10	n/a	n/a	60.10	n/a
S98T002818	D	Formate by Ion Chromatograph	ug/mL	92.31	<2.10e-02	1.07e+03	1.08e+03	1.08e+03	0.93	n/a	131.5	n/a
S98T002818	D	Fluoride-IC-Dionex 4000/4500	ug/mL	100.9	<1.20e-02	<61.81	<61.81	<61.81	n/a	n/a	61.81	n/a
S98T002818	D	Chloride-IC-Dionex 4000/4500	ug/mL	103.1	<1.00e-02	5.04e+03	5.23e+03	5.13e+03	3.70	n/a	87.57	n/a
S98T002818	D	Nitrate-IC-Dionex 4000/4500	ug/mL	104.1	<1.00e-01	7.09e+04	7.34e+04	7.21e+04	3.47	n/a	556.3	n/a
S98T002818	D	Bromide by Ion Chromatograph	ug/mL	96.98	<1.25e-01	<6.44e+02	<6.44e+02	<6.44e+02	n/a	n/a	643.9	n/a
S98T002818	D	Nitrate by IC-Dionex 4000/4500	ug/mL	98.14	<1.20e-01	1.88e+03	1.90e+03	1.89e+03	1.06	n/a	716.0	n/a
S98T002818	D	Phosphate by IC-Dionex 4000/4500	ug/mL	100.5	<1.20e-01	1.53e+03	<6.18e+02	<6.18e+02	3.51	n/a	618.1	n/a
S98T002818	D	Sulfate by IC-Dionex 4000/4500	ug/mL	101.4	<1.38e-01	2.32e+03	2.24e+03	2.28e+03	3.51	n/a	710.8	n/a
S98T002818	D	Oxalate by IC-Dionex 4000/4500	ug/mL	106.6	<1.05e-01	<5.41e+02	<5.41e+02	<5.41e+02	n/a	n/a	540.9	n/a
S98T002818	D	Cesium-137 by GEA	ug/mL	99.69	<1.52e-03	2.18e+02	220.0	219.0	0.91	n/a	n/a	0.110
S98T002826	D	Chromium (VI) by Spec.	ug/mL	100.4	<2.50e-03	1.58e+03	1.55e+03	1.55e+03	1.30	n/a	1.00e-02	n/a
S98T002826	D	1.129 by ICP/MS Acid Addition	ug/mL	104.7	<1.20e-04	<4.860	<4.860	<4.860	n/a	n/a	4.860	n/a
S98T002826	D	GM- by Pot. Titration	ug/mL	93.33	<1.00e-04	5.14e+04	4.99e+04	5.06e+04	2.96	n/a	2.50e+03	n/a
S98T002826	D	TOC by Persulfate/Coulometry	ug/mL	93.33	<1.20e-03	2.67e+03	2.67e+03	2.60e+03	4.92	n/a	40.00	n/a
S98T002826	D	Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	n/a	n/a	4.86e-01	n/a
S98T002826	D	Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	n/a	n/a	4.86e-01	n/a
S98T002826	D	Uranium-235 by ICP/MS Acid Add	ug/mL	101.4	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	n/a	n/a	4.86e-01	n/a
S98T002826	D	Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<1.60e-05	<6.48e-01	<6.48e-01	<6.48e-01	n/a	n/a	6.48e-01	n/a
S98T002826	D	Uranium-238 by ICP/MS Acid Add	ug/mL	104.0	<1.20e-05	5.285	5.290	5.287	0.00	n/a	4.86e-01	n/a
S98T002826	D	Tc 99 by ICP/MS Acid Addition	ug/mL	85.37	<2.00e-06	5.403	5.400	5.402	0.00	n/a	8.10e-02	n/a

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002826	D	Sn 126	by ICP/MS Acid Addition	ug/mL								
S981002826	A	Neptunium-237	by ICP/MS	ug/mL	n/a	<1.50e-04	< 6.07e-01	<6.08e-01	n/a	n/a	6.07e-01	n/a
S981002826	A	Neptunium-237	by ICP/MS	ug/mL	119.6	<4.67e-06	<1.89e-01	<1.89e-01	n/a	n/a	1.89e-01	n/a
S981002826	A	Plutonium-239	by ICP/MS	ug/mL	104.2	<3.47e-06	<1.41e-01	<1.41e-01	n/a	n/a	1.41e-01	n/a
S981002826	A	Plutonium-240	by ICP/MS	ug/mL	n/a	<3.55e-06	<1.44e-01	<1.44e-01	n/a	n/a	1.44e-01	n/a
S981002826	A	Plutonium-242	by ICP/MS	ug/mL	n/a	<2.78e-06	<1.13e-01	<1.13e-01	n/a	n/a	1.13e-01	n/a
S981002826	A	Pu/Am-241	by ICP/MS	ug/mL	109.7	<3.71e-06	<1.50e-01	<1.50e-01	n/a	n/a	1.50e-01	n/a

Table 1. Data Summary Report.
BT-105 (2)

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S9810022811	D	DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S9810022811	D	DSC Exotherm on Perkin Elmer	Joules/g	97.96	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S9810022811	D	Ammonia by ISE-Std Additions	ug/mL	97.00	<25.00	2.00e-03	157.0	178.5	24.1	113.0	25.00	n/a
S9810022811	D	Specific Gravity	Sp.G.	99.05	n/a	1.428	1.428	1.429	0.07	n/a	1.00e-03	n/a
S9810022811	D	% Water by TGA on Perkin Elmer	%	99.29	n/a	48.49	52.44	50.56	7.42	n/a	n/a	n/a
S9810022811	D	Alpha in Liquid Samples	uCi/mL	97.78	<1.26e-02	<1.03e-02	<1.37E-2	n/a	n/a	n/a	1.90e-02	5.00E+02
S9810022811	D	Strontium-89/90 High Level	uCi/mL	101.9	<2.00e-03	1.32e-04	1.47E-4	1.40E-01	10.8	n/a	2.00e-03	5.29E+00
S9810022811	D	Pu-239/240 by TRU-SPEC Resin	uCi/mL	105.4	<3.35e-04	3.47E-04	4.85E-04	4.85E-04	0.00	-5.84e-1	3.47E-04	1.00E+02
S9810022811	D	Aluminum-TC-Acid Dtl.	ug/mL	97.20	<3.00e-03	4.86e-04	35.50	35.65	0.84	99.50	30.10	n/a
S9810022811	D	Boron-TC-Acid Dtl.	ug/mL	102.3	<3.00e-03	35.50	35.80	35.65	0.84	99.50	30.10	n/a
S9810022811	D	Barium-TC-Acid Dtl.	ug/mL	96.80	<3.00e-03	<30.10	<3.01e1	n/a	n/a	n/a	91.90	30.10
S9810022811	D	Bismuth-TC-Acid Dtl.	ug/mL	100.6	<1.00e-01	<40.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S9810022811	D	Calcium-TC-Acid Dtl.	ug/mL	103.0	<1.00e-01	<40.10	<6.01e1	n/a	n/a	n/a	105.0	60.10
S9810022811	D	Chromium-TC-Acid Dtl.	ug/mL	99.40	<1.00e-02	1.59e+03	1.57e+03	0.63	94.00	6.010	n/a	n/a
S9810022811	D	Iron-TC-Acid Dtl.	ug/mL	95.60	<1.00e-02	32.40	32.80	32.60	1.23	93.30	30.10	n/a
S9810022811	D	Potassium-TC-Acid Dtl.	ug/mL	99.60	<1.00e-01	4.85e+03	4.91e+03	4.88e+03	1.23	73.80	300.0	n/a
S9810022811	D	Lithium-TC-Acid Dtl.	ug/mL	99.60	<1.00e-02	<6.010	<6.010	n/a	n/a	n/a	93.80	6.010
S9810022811	D	Manganese-TC-Acid Dtl.	ug/mL	98.40	<1.00e-02	<6.010	<6.010	n/a	n/a	n/a	95.00	6.010
S9810022811	D	Sodium-TC-Acid Dtl.	ug/mL	98.40	<1.00e-02	2.13e+05	2.15e+05	2.14e+05	0.93	-7.550e2	60.10	n/a
S9810022811	D	Nickel-TC-Acid Dtl.	ug/mL	103.0	<1.00e-02	<12.00	<1.20e1	n/a	n/a	n/a	96.00	12.00
S9810022811	D	Uranium-TC-Acid Dtl.	ug/mL	102.2	<5.00e-02	2.53e+02	256.0	254.5	1.18	105.0	30.10	n/a
S9810022811	D	Strontium-TC-Acid Dtl.	ug/mL	99.60	<1.00e-01	<3.01e2	<3.01e2	n/a	n/a	n/a	94.50	300.0
S9810022811	D	Uranium-TC-Acid Dtl.	ug/mL	97.20	<1.00e-02	<6.010	<6.010	n/a	n/a	n/a	94.40	6.010
S9810022811	D	Zinc-TC-Acid Dtl.	ug/mL	97.20	<1.00e-02	<6.010	<6.010	n/a	n/a	n/a	96.50	6.010
S9810022811	D	Zirconium-TC-Acid Dtl.	ug/mL	100.0	<2.10e-02	1.29e+03	1.38e+03	1.34e+03	6.74	112.6	37.57	n/a
S9810022811	D	Formate by Ion Chromatograph	ug/mL	103.0	<1.20e-02	<61.81	<6.18e1	n/a	n/a	n/a	61.81	n/a
S9810022811	D	Fluoride-TC-Bioxex 4000/4500	ug/mL	100.9	<1.20e-02	5.19e+03	5.19e+03	0.97	99.63	87.57	n/a	n/a
S9810022811	D	Chloride-TC-Bioxex 4000/4500	ug/mL	100.9	<1.08e-01	7.45e+04	7.45e+04	0.00	102.8	556.3	n/a	n/a
S9810022811	D	Nitrate by Ion Chromatograph	ug/mL	96.98	<1.08e-01	<6.44e+02	<6.44e2	n/a	n/a	n/a	91.98	643.9
S9810022811	D	Bromide by Ion Chromatograph	ug/mL	98.12	<1.39e-01	1.87e+05	1.86e+05	0.00	91.98	716.0	n/a	n/a
S9810022811	D	Nitrate by IC-Bioxex 4000/4500	ug/mL	100.5	<1.20e-01	8.70e+02	1.07e+03	969.9	20.6	92.50	618.1	n/a
S9810022811	D	Phosphate-TC-Bioxex 4000/4500	ug/mL	100.2	<1.38e-01	2.13e+03	2.07e+03	2.86	99.22	710.8	n/a	n/a
S9810022811	D	Sulfate by IC-Bioxex 4000/4500	ug/mL	103.6	<1.05e-01	<5.41e+02	<5.41e2	n/a	n/a	n/a	99.82	540.9
S9810022811	D	Oxalate-TC-Bioxex 4000/4500	ug/mL	99.69	<1.52e-03	2.17e+02	220.0	218.5	1.37	n/a	n/a	0.110
S9810022811	D	Cesium-137 by GEA	uCi/mL	100.0	<2.58e-01	1.03e+03	1.04e+03	0.97	98.00	1.00e-02	n/a	n/a
S9810022811	D	Chemium (VI) by Spec.	ug/mL	104.7	<1.20e-04	<4.860	<4.860	n/a	n/a	n/a	4.860	n/a
S9810022811	D	OH- by Pot. Titration	ug/mL	100.0	<1.20e-04	5.46e+04	5.35e+04	2.04	n/a	2.50e+03	n/a	n/a
S9810022811	D	TDC by Persulfate/Coulometry	ug/mL	94.67	8.700	2.76e+03	2.75e+03	0.73	97.30	40.00	n/a	n/a
S9810022811	D	Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<5.07e-05	<2.053	<2.05e0	n/a	n/a	2.053	n/a	n/a
S9810022811	D	Uranium-234 by ICP/MS Acid Add	ug/mL	102.1	<4.10e-05	<1.659	<1.66e0	n/a	n/a	1.659	n/a	n/a
S9810022811	D	Uranium-235 by ICP/MS Acid Add	ug/mL	102.1	<4.35e-05	<1.845	<1.84e0	n/a	n/a	1.845	n/a	n/a
S9810022811	D	Uranium-236 by ICP/MS Acid Add	ug/mL	105.0	<3.37e-05	<1.448	<1.45e0	n/a	n/a	1.448	n/a	n/a
S9810022811	D	Uranium-238 by ICP/MS Acid Add	ug/mL	105.0	<5.86e-05	<2.373	<2.37e0	n/a	n/a	100.0	2.373	n/a
S9810022811	D	Tc 99 by ICP/MS Acid Addition	ug/mL	81.73	<2.00e-06	4.192	4.290	4.241	2.36	79.45	8.10e-02	n/a

Sample#	R/L#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002827	D	Sn 126 by ICP/MS	ug/mL	n/a	<1.50e-04	< 6.075	<6.0860	n/a	n/a	n/a	6.075	n/a
S98T002827	A	Se/Unlamin 237 by ICP/MS	ug/mL	134.1	n/a	<6.53e-01	<6.53e-01	n/a	n/a	147.0	6.53e-01	n/a
S98T002827	A	Plutonium 239 by ICP/MS	ug/mL	139.0	<1.75e-03	<7.09e-01	<7.09e-01	n/a	n/a	129.0	7.09e-01	n/a
S98T002827	A	Plutonium 240 by ICP/MS	ug/mL	n/a	<1.52e-03	<6.14e-01	<6.14e-01	n/a	n/a	n/a	4.14e-01	n/a
S98T002827	A	Plutonium 241 by ICP/MS	ug/mL	n/a	<1.52e-03	<6.08e-01	<6.08e-01	n/a	n/a	n/a	6.08e-01	n/a
S98T002827	A	Plutonium 241 by ICP/MS	ug/mL	n/a	<1.58e-03	<6.41e-01	<6.41e-01	n/a	n/a	n/a	6.41e-01	n/a

CORE NUMBER: 24LR
SEGMENT #: 8R

Table 1. Data Summary Report.
81-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002838		Bulk Density of Sample	g/ml	N/A	N/A	1.820	N/A	N/A	N/A	N/A	5.00e-01	N/A
S98T002840		DSC Exotherm DRY Calculated	Joules/g DRY	N/A	N/A	12.28	N/A	38.8	N/A	N/A	N/A	N/A
S98T002840		DSC Exotherm on Perkin Elmer	Joules/g	98.84	N/A	25.88	33.80	22.3	N/A	N/A	N/A	N/A
S98T002840		% Water by TGA on Perkin Elmer	%	99.17	N/A	26.36	33.86	26.3	N/A	N/A	N/A	N/A
S98T002840		TIC by Acid/Coulometry	ug/g	99.50	1.00	2.42e+03	2.98e+03	27.1	50.6	188.0	5.000	N/A
S98T002840		TOC by Persulfate/Coulometry	ug/g	91.67	1.00	2.49e+03	2.16e+03	18.7	18.7	188.0	5.000	N/A
S98T002840		F Strontium - ICP-Fusion	ug/g	97.11	6.70e-05	36.30	36.45	0.82	16.6	99.80	1.00e-02	7.70e-01
S98T002844		F Aluminum - ICP-Fusion	ug/g	99.20	5.00e-02	1.38e+03	1.35e+03	16.6	99.80	1.03e+03	N/A	N/A
S98T002844		F Boron - ICP-Fusion	ug/g	104.6	5.00e-02	1.03e+03	1.03e+03	N/A	10.0	99.80	1.03e+03	N/A
S98T002844		F Barium - ICP-Fusion	ug/g	99.00	5.00e-02	1.03e+03	1.03e+03	N/A	10.0	99.80	1.03e+03	N/A
S98T002844		F Bismuth - ICP-Fusion	ug/g	101.4	5.00e-02	1.03e+03	1.03e+03	N/A	10.0	99.80	1.03e+03	N/A
S98T002844		F Calcium - ICP-Fusion	ug/g	104.0	5.00e-02	3.32e+03	2.70e+03	17.8	99.20	2.00e+03	N/A	N/A
S98T002844		F Chromium - ICP-Fusion	ug/g	101.2	5.00e-02	8.59e+02	8.95e+02	8.7	99.20	2.00e+03	N/A	N/A
S98T002844		F Iron - ICP-Fusion	ug/g	97.80	5.00e-02	1.78e+03	1.90e+03	6.52	95.60	2.00e+03	N/A	N/A
S98T002844		F Lithium - ICP-Fusion	ug/g	103.2	5.00e-02	2.07e+02	2.06e+02	N/A	17.8	97.60	2.00e+03	N/A
S98T002844		F Manganese - ICP-Fusion	ug/g	100.0	5.00e-02	2.07e+02	2.06e+02	N/A	17.8	97.60	2.00e+03	N/A
S98T002844		F Sodium - ICP-Fusion	ug/g	102.6	5.00e-02	2.12e+03	2.12e+03	0.08	16.2	0.0	2.00e+03	N/A
S98T002844		F Nickel - ICP-Fusion	ug/g	100.2	1.68e-01	5.69e+03	5.39e+03	5.44e+03	1.80	99.80	1.03e+03	N/A
S98T002844		F Silicon - ICP-Fusion	ug/g	104.2	5.00e-02	1.74e+03	2.18e+03	1.94e+03	22.4	103.0	1.03e+03	N/A
S98T002844		F Uranium - ICP-Fusion	ug/g	98.90	5.00e-02	1.03e+02	1.03e+02	N/A	17.8	97.50	1.03e+03	N/A
S98T002844		F Zinc - ICP-Fusion	ug/g	95.40	5.00e-02	2.07e+02	2.06e+02	N/A	17.8	94.60	2.00e+03	N/A
S98T002844		F Zirconium - ICP-Fusion	ug/g	98.60	5.00e-02	2.07e+02	2.06e+02	N/A	17.8	95.90	2.00e+03	N/A
S98T002844		F Cesium-137 by BEA	ug/g	102.3	1.87e-01	65.48	69.20	67.34	5.49	N/A	207.0	N/A
S98T002844		F Alpha of Digested Solid	ug/g	89.44	2.45e-01	2.78e-01	4.34e-01	N/A	87.65	3.38e-01	5.00e+02	N/A
S98T002844		F Alpha of ICP MS Acid/g LA-505159	ug/g	109.0	4.120e-03	4.918	44.87e-01	N/A	N/A	N/A	4.918	N/A
S98T002844		F Uranium-233 by ICP MS Acid/g LA-505159	ug/g	N/A	4.32e-04	44.92e-01	44.87e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002844		F Uranium-235 by ICP MS Acid/g LA-505159	ug/g	N/A	4.32e-04	44.92e-01	44.87e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002844		F Uranium-235 by ICP MS Acid/g LA-505159	ug/g	98.59	4.32e-04	15.00	15.40	15.20	2.63	97.7	4.92e-01	N/A
S98T002844		F Uranium-238 by ICP MS Acid/g LA-505159	ug/g	N/A	4.32e-04	44.92e-01	44.87e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002844		F Uranium-238 by ICP MS Acid/g LA-505159	ug/g	94.50	4.32e-04	2.30e+03	2.34e+03	2.32e+03	1.72	87.00	4.92e-01	N/A
S98T002844		F Tc99 by ICP MS Acid/g LA-505159	ug/g	93.66	4.91e-04	1.701	1.650	1.675	2.99	N/A	8.00e-02	N/A
S98T002844		F Sn126 ICP MS Acid/g LA-505159	ug/g	N/A	5.00e-03	6.148	6.0960	N/A	N/A	N/A	6.149	N/A
S98T002844		F Neptunium-237 by ICP MS	ug/g	118.9	2.31e-05	48.59e-02	48.51e-02	N/A	N/A	113.4	8.60e-02	N/A
S98T002844		F Plutonium-239 by ICP MS	ug/g	99.60	4.01e-05	1.49e-01	1.48e-01	N/A	N/A	97.61	1.49e-01	N/A
S98T002844		F Plutonium-240 by ICP MS	ug/g	N/A	4.12e-05	44.10e-02	44.12e-02	N/A	N/A	N/A	4.20e-02	N/A
S98T002844		F Plutonium-242 by ICP MS	ug/g	N/A	4.35e-05	1.29e-01	1.27e-01	N/A	N/A	N/A	1.29e-01	N/A
S98T002844		F Pu/Am-241 by ICP MS	ug/g	103.9	4.35e-05	46.88e-02	46.81e-02	N/A	N/A	101.9	6.90e-02	N/A
S98T002844		F Aluminum - ICP-Acid Digest	ug/g	96.40	7.27e-01	1.53e+04	1.60e+04	1.56e+04	4.47	-3.06e3	51.00	N/A
S98T002844		F Boron - ICP-Acid Digest	ug/g	107.4	4.27e-01	88.00	105.0	96.50	17.6	105.8	51.00	N/A
S98T002844		F Barium - ICP-Acid Digest	ug/g	93.00	5.00e-02	5.00e-02	5.05e-02	N/A	N/A	97.80	51.00	N/A
S98T002844		F Bismuth - ICP-Acid Digest	ug/g	91.80	5.00e-02	1.74e+02	1.51e+02	162.0	13.6	99.00	102.0	N/A
S98T002844		F Calcium - ICP-Acid Digest	ug/g	105.0	1.57e-01	2.18e+03	3.39e+03	2.78e+03	43.4	112.2	102.0	N/A
S98T002844		F Chromium - ICP-Acid Digest	ug/g	94.20	5.00e-02	6.63e+02	8.73e+02	868.0	1.15	94.00	10.20	N/A
S98T002844		F Iron - ICP-Acid Digest	ug/g	91.20	5.00e-02	2.00e+03	2.05e+03	2.02e+03	2.47	87.60	51.00	N/A

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	Rpp % Spk Rec %	Det Limit	Count	Err%
S98T002846	A	Potassium - ICP-Acid Digest	ug/g	96.20	<5.00e-01	1.59e+03	1.28e+03	1.44e+03	21.6	60.80	510.0	N/A
S98T002846	A	Lithium - ICP-Acid Digest	ug/g	97.60	<1.00e-02	< 10.20	< 10.20	< 10.20	99.20	99.20	10.20	N/A
S98T002846	A	Manganese - ICP-Acid Digest	ug/g	92.20	<1.00e-02	37.80	38.40	38.10	1.35	98.20	10.20	N/A
S98T002846	A	Sodium - ICP-Acid Digest	ug/g	129.2	1.150	2.24e+05	2.21e+05	2.23e+05	1.35	54.8	102.0	N/A
S98T002846	A	Nickel - ICP-Acid Digest	ug/g	91.80	<2.00e-02	8.17e+02	8.42	8.29	3.01	97.00	20.40	N/A
S98T002846	A	Silicon - ICP-Acid Digest	ug/g	85.80	2.130	8.21e+02	1.15e+03	985.5	33.4	129.4	51.00	N/A
S98T002846	A	Uranium - ICP-Acid Digest	ug/g	93.40	<5.00e-01	2.28e+03	2.31e+03	2.30e+03	1.31	89.00	510.0	N/A
S98T002846	A	Zinc - ICP-Acid Digest	ug/g	84.60	<1.00e-02	12.10	10.80	11.45	11.45	97.60	10.20	N/A
S98T002846	A	Zirconium - ICP-Acid Digest	ug/g	93.00	<1.00e-02	< 10.20	< 10.20	< 10.20	N/A	98.00	10.20	N/A
S98T002846	A	Formate by IC-Dioxin 4000/4500	ug/g	91.40	<6.20e-02	<2.55e+02	<2.55e+02	N/A	N/A	83.59	254.9	N/A
S98T002846	A	Fluoride-IC-Dioxin 4000/4500	ug/g	98.29	<1.20e-02	6.74e+02	1.69e+03	1.18e+03	86.0	71.92	237.2	N/A
S98T002846	A	Chloride-IC-Dioxin 4000/4500	ug/g	102.7	5.40e-02	1.13e+03	1.23e+03	1.18e+03	8.47	71.92	336.1	N/A
S98T002846	A	Nitrite-IC - Dioxin 4000/4500	ug/g	106.2	<1.00e-01	1.42e+04	1.52e+04	1.47e+04	6.80	70.91	2.14e+03	N/A
S98T002846	A	Bromide by Ion Chromatograph	ug/g	101.2	<1.25e-01	<2.47e+03	<2.47e+03	N/A	N/A	68.06	2.47e+03	N/A
S98T002846	A	Nitrate by IC-Dioxin 4000/4500	ug/g	93.49	<1.20e-01	5.32e+05	5.04e+05	5.18e+05	5.41	72.43	2.79e+03	N/A
S98T002846	A	Phosphate by IC-Dioxin 4000/4500	ug/g	102.2	<1.20e-01	5.49e+03	8.15e+03	6.82e+03	39.0	68.93	3.37e+03	N/A
S98T002846	A	Sulfate by IC-Dioxin 4000/4500	ug/g	101.9	<1.38e-01	2.88e+03	3.58e+03	3.23e+03	21.7	68.54	2.79e+03	N/A
S98T002846	A	Oxalate by IC-Dioxin 4000/4500	ug/g	101.7	<1.05e-01	3.04e+03	3.35e+03	3.20e+03	9.70	70.00	2.08e+03	N/A

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	Rpp % Spk Rec %	Det Limit	Count	Err%
S98T002812	D	DSC Exotherm Dry	Joules/g Dry	N/A	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S98T002812	D	DSC Exotherm on Perkin Elmer	Joules/g	98.84	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S98T002812	D	Ammonia by ISE-Std Additions	ug/mL	113.8	<5.000	63.30	220.0	141.7	111	N/A	55.00	N/A
S98T002812	D	Specific Gravity	Sp. G.	99.06	N/A	1.434	1.417	1.425	1.19	N/A	1.00e-03	N/A
S98T002812	D	% Water by TGA on Perkin Elmer	%	99.29	N/A	52.20	52.10	52.15	0.19	N/A	N/A	N/A
S98T002812	D	Alpha In Liquid Samples	uCi/mL	97.78	<1.26e-02	<1.49e-02	<1.37E-2	N/A	N/A	1.90e-02	5.00E+02	N/A
S98T002820	D	Strontium-89/90 by High Level	uCi/mL	101.9	<2.00e-03	8.35e-02	8.16e-02	8.26e-02	2.30	N/A	2.00e-03	6.74E+00
S98T002820	D	Pu-239/240 by TRU-SPEC Resin	uCi/mL	105.4	<3.35e-04	<3.43e-04	<3.69E-4	N/A	N/A	3.43e-04	8.38E+00	N/A
S98T002820	D	Aluminum-IC-Acid Dtl.	ug/mL	97.80	<5.00e-02	4.96e+04	4.96e+04	4.96e+04	0.20	N/A	30.10	N/A
S98T002820	D	Boron-IC-Acid Dtl.	ug/mL	101.2	<5.00e-02	< 30.10	< 30.10	N/A	N/A	N/A	30.10	N/A
S98T002820	D	Barium-IC-Acid Dtl.	ug/mL	96.80	<5.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	30.10	N/A
S98T002820	D	Bismuth-IC-Acid Dtl.	ug/mL	100.2	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002820	D	Calcium-IC-Acid Dtl.	ug/mL	103.0	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002820	D	Chromium-IC-Acid Dtl.	ug/mL	99.80	<1.00e-02	1.58e+03	1.58e+03	1.58e+03	0.00	N/A	6.010	N/A
S98T002820	D	Iron-IC-Acid Dtl.	ug/mL	96.40	<5.00e-02	36.40	36.10	36.25	0.83	N/A	30.10	N/A
S98T002820	D	Potassium-IC-Acid Dtl.	ug/mL	97.40	<5.00e-01	5.05e+03	4.96e+03	5.00e+03	1.80	N/A	300.0	N/A
S98T002820	D	Lithium-IC-Acid Dtl.	ug/mL	99.80	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002820	D	Manganese-IC-Acid Dtl.	ug/mL	98.60	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002820	D	Sodium-IC-Acid Dtl.	ug/mL	99.20	<1.00e-01	2.21e+05	2.17e+05	2.19e+05	1.83	N/A	60.10	N/A
S98T002820	D	Nickel-IC-Acid Dtl.	ug/mL	99.20	<1.00e-02	< 12.00	< 12.0e1	N/A	N/A	N/A	12.00	N/A
S98T002820	D	Silicon-IC-Acid Dtl.	ug/mL	103.8	<5.00e-02	2.23e+02	226.0	224.5	1.34	N/A	30.10	N/A
S98T002820	D	Uranium-IC-Acid Dtl.	ug/mL	96.20	<5.00e-02	<3.01e02	<3.01e2	N/A	N/A	N/A	300.0	N/A
S98T002820	D	Zinc-IC-Acid Dtl.	ug/mL	94.60	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002820	D	Zirconium-IC-Acid Dtl.	ug/mL	96.80	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002820	D	Formate by Ion Chromatograph	ug/mL	100.0	<2.10e-02	1.31e+03	1.28e+03	1.30e+03	2.32	N/A	37.57	N/A
S98T002820	D	Fluoride-IC-Dioxin 4000/4500	ug/mL	101.1	<1.20e-02	< 61.81	<6.18e1	N/A	N/A	N/A	61.81	N/A
S98T002820	D	Chloride-IC-Dioxin 4000/4500	ug/mL	105.2	<1.70e-02	5.31e+03	5.36e+03	5.34e+03	0.94	N/A	87.57	N/A
S98T002820	D	Nitrite-IC - Dioxin 4000/4500	ug/mL	102.5	<1.08e-01	7.51e+04	7.46e+04	7.49e+04	0.67	N/A	556.3	N/A
S98T002820	D	Bromide by Ion Chromatograph	ug/mL	99.52	<1.25e-01	<6.44e+02	<6.44e2	N/A	N/A	N/A	643.9	N/A

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002820		Nitrate by IC-Dionex 4000/4500	ug/mL	99.57	<1.39e-01	1.79e+05	1.79e+05	1.79e+05	0.00	n/a	716.0	n/a
S981002820		Phosphate-IC-Dionex 4000/4500	ug/mL	104.0	<1.20e-01	1.42e+03	1.39e+03	1.41e+03	2.14	n/a	618.1	n/a
S981002820		Sulfate by IC-Dionex 4000/4500	ug/mL	101.4	<1.39e-01	2.60e+03	2.54e+03	2.57e+03	2.33	n/a	710.8	n/a
S981002820		Oxalate-IC-Dionex 4000/4500	ug/mL	106.8	<1.05e-01	9.88e+02	587.0	787.4	50.9	n/a	540.9	n/a
S981002820		Cesium-137 by Gea	uCi/mL	99.69	<1.52e-03	2.53e+02	251.0	252.0	0.79	n/a	n/a	0.190
S981002828		Chromium (VI) by Spec.	ug/mL	100.0	<2.58e-01	1.35e+03	1.37e+03	1.36e+03	1.47	n/a	1.00e-02	n/a
S981002828	D	1.12 by IC/MS Acid Addition	ug/mL	104.7	<1.20e-04	< 4.860	<4.860	n/a	n/a	n/a	4.860	n/a
S981002828		OH- by Pot. Titration	ug/mL	100.0	n/a	6.19e+04	5.74e+04	5.94e+04	6.90	n/a	2.50e+03	n/a
S981002828		TOL by Persulfate/Coulometry	ug/mL	94.67	8.700	2.67e+03	2.80e+03	2.74e+03	4.75	n/a	40.00	n/a
S981002828	D	Uranium-233 by IC/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002828	D	Uranium-234 by IC/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002828	D	Uranium-232 by IC/MS Acid Add	ug/mL	101.4	<1.20e-05	<4.86e-01	<4.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002828	D	Uranium-235 by IC/MS Acid Add	ug/mL	n/a	<1.60e-05	<6.48e-01	<6.48e-01	n/a	n/a	n/a	6.48e-01	n/a
S981002828	D	Uranium-238 by IC/MS Acid Add	ug/mL	104.0	<1.20e-05	<4.86e-01	<4.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002828	D	IC 99 by IC/MS Acid Addition	ug/mL	85.37	<2.00e-06	< 5.172	5.100	5.136	1.36	n/a	8.10e-02	n/a
S981002828	D	Sn 126 by IC/MS Acid Addition	ug/mL	n/a	<1.50e-04	< 6.075	<6.0860	n/a	n/a	n/a	6.075	n/a
S981002828	A	Neptunium-237 by IC/MS	ug/mL	119.6	<4.67e-06	<1.89e-01	<1.89e-01	n/a	n/a	n/a	1.89e-01	n/a
S981002828	A	Plutonium-239 by IC/MS	ug/mL	104.2	<3.47e-06	<1.41e-01	<1.41e-01	n/a	n/a	n/a	1.41e-01	n/a
S981002828	A	Plutonium-240 by IC/MS	ug/mL	n/a	<3.55e-06	<1.44e-01	<1.44e-01	n/a	n/a	n/a	1.44e-01	n/a
S981002828	A	Plutonium-242 by IC/MS	ug/mL	109.4	<2.78e-06	<1.13e-01	<1.13e-01	n/a	n/a	n/a	1.13e-01	n/a
S981002828	A	Pu/Am-241 by IC/MS	ug/mL	109.7	<3.71e-06	<1.50e-01	<1.50e-01	n/a	n/a	n/a	1.50e-01	n/a

CORE NUMBER: 246R
SEGMENT #: 9

Table 1. Data Summary Report.
Bv-105 (Z)

SEGMENT PORTION: L Lower Half of Segment

Sample#	RAD Analyte	Unit	Standard %	Blank	Result	Duplicates	Average	Rep. % Spk Rec %	Det Limit	Count Err%
S98T002839	Bulk Density of Sample	g/cm ³			1.770					
S98T002840	ISC Exotherm Dry, Calculated	Joules/g Dry			36.09	58.33	47.21		5.00e+01	
S98T002841	ISC Exotherm on Perkin Elmer	Joules/g Dry	98.84		28.56	46.16	37.36	47.1		
S98T002842	Water by IGA on Perkin Elmer	%	99.29		26.21	15.01	20.86	56.1		
S98T002843	Water by IGA on Perkin Elmer	%	95.56	5.90	4.39+03	4.58+03	4.48+03	4.24		
S98T002844	IC by Act Coulometry	ug/g	95.42	5.200	8.26+03	1.05e+04	9.62+03	28.7		
S98T002845	Strontium-89 by Radiometry	ug/g	10.97	< 1.0e-01	2.80+02	293.0	286.5	4.54		
S98T002846	Aluminum-27 by Fusion	ug/g	96.40	< 5.0e-02	7.71e+04	1.80e+04	1.76e+04	3.97		
S98T002847	Boron-10 by Fusion	ug/g	103.0	< 5.0e-02	< 1.0e+03	< 1.0e+03				
S98T002848	Boron-11 by Fusion	ug/g	96.40	< 5.0e-02	< 1.0e+03	< 1.0e+03				
S98T002849	Bismuth-210 by Fusion	ug/g	109.4	< 1.0e-01	< 2.0e+03	< 2.0e+03				
S98T002850	Calcium-44 by Fusion	ug/g	103.0	< 1.0e-01	0.89e+04	1.13e+04	1.10e+04	4.52		
S98T002851	Chromium-52 by Fusion	ug/g	99.60	< 5.0e-02	1.06e+03	1.36e+04	1.32e+04	5.28		
S98T002852	Iron-56 by Fusion	ug/g	96.40	< 1.0e-02	< 2.0e+02	< 2.0e+02				
S98T002853	Lithium-7 by Fusion	ug/g	98.20	< 1.0e-02	1.75e+05	1.76e+05	1.76e+05	0.57		
S98T002854	Manganese-55 by Fusion	ug/g	98.20	< 1.0e-01	1.42e+03	1.59e+03	1.50e+03	11.3		
S98T002855	Sodium-23 by Fusion	ug/g	99.80	< 1.0e-01	3.09e+03	3.04e+03	3.06e+03	1.63		
S98T002856	Silicon-28 by Fusion	ug/g	104.0	< 5.0e-02	1.44e+04	1.51e+04	1.48e+04	4.75		
S98T002857	Zinc-66 by Fusion	ug/g	96.70	< 5.0e-01	< 2.0e+02	< 2.0e+02				
S98T002858	Zinc-68 by Fusion	ug/g	97.20	< 1.0e-02	< 2.0e+02	< 2.0e+02				
S98T002859	Zinc-70 by Fusion	ug/g	103.4	< 1.75e-01	1.06e+02	1.10e+02	108.2	3.70		
S98T002860	Cadmium-112 by Fusion	ug/g	95.56	< 2.7e-01	< 3.27e+01	2.95e+01				
S98T002861	Alpha of Digested Solid	ug/g	109.0	< 1.20e-03	< 5.0e+02	44.60e2				
S98T002862	Uranium-235 by ICP-MS	ug/g								
S98T002863	Uranium-238 by ICP-MS	ug/g								
S98T002864	Uranium-235 by ICP-MS	ug/g								
S98T002865	Uranium-238 by ICP-MS	ug/g								
S98T002866	Uranium-235 by ICP-MS	ug/g	105.6	< 1.20e-04	< 5.008	44.55e0				
S98T002867	Uranium-238 by ICP-MS	ug/g	106.0	< 1.60e-04	< 6.678	46.07e0	120.1	8.33		
S98T002868	Uranium-235 by ICP-MS	ug/g								
S98T002869	Uranium-238 by ICP-MS	ug/g	91.22	< 3.66e-04	< 1.77e+04	1.92e+04	1.84e+04	8.13		
S98T002870	Thorium-232 by ICP-MS	ug/g	114.2	< 1.65e-03	< 6.32e+02	< 5.79e2				
S98T002871	Plutonium-239 by ICP-MS	ug/g	98.51	< 3.56e-05	< 5.532	< 5.038				
S98T002872	Plutonium-240 by ICP-MS	ug/g								
S98T002873	Plutonium-242 by ICP-MS	ug/g								
S98T002874	Plutonium-244 by ICP-MS	ug/g	102.3	< 1.69e-05	< 2.631	< 2.390				
S98T002875	Aluminum-27 by Fusion	ug/g	96.40	< 1.97e-05	< 3.063	< 2.780				
S98T002876	Boron-10 by Fusion	ug/g	104.0	< 5.29e-02	1.82e+04	2.17e+04	2.00e+04	17.5		
S98T002877	Boron-11 by Fusion	ug/g	91.40	< 1.00e-02	1.74e+02	207.0	190.5	17.3		
S98T002878	Bismuth-210 by Fusion	ug/g	97.20	< 1.92e-01	1.27e+03	1.34e+03	1.30e+03	5.36		
S98T002879	Calcium-44 by Fusion	ug/g	93.00	< 1.00e-02	1.01e+03	3.15e+04	2.08e+04	103		
S98T002880	Chromium-52 by Fusion	ug/g	90.60	< 5.00e-02	1.24e+04	1.31e+04	1.28e+04	5.49		
S98T002881	Iron-56 by Fusion	ug/g								

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002847	A	Potassium - ICP-Acid Digest	ug/g	85.60	<5.00e-01	1.44e+03	1.38e+03	1.26e+03	n/a	n/a	312.0	n/a
S981002847	A	Lithium - ICP-Acid Digest	ug/g	99.00	<1.00e-02	< 6.230	< 6.660	n/a	n/a	n/a	6.240	n/a
S981002847	A	Manganese - ICP-Acid Digest	ug/g	91.20	<1.00e-02	1.74e+02	178.0	176.0	2.27	n/a	6.240	n/a
S981002847	A	Sodium - ICP-Acid Digest	ug/g	115.0	8.55e-01	1.80e+05	1.70e+05	1.75e+05	5.71	n/a	62.40	n/a
S981002847	A	Nickel - ICP-Acid Digest	ug/g	90.40	<2.00e-02	5.26e+03	5.70e+03	5.48e+03	8.03	n/a	12.50	n/a
S981002847	A	Silicon - ICP-Acid Digest	ug/g	154.2	6.95e-01	3.21e+03	6.41e+03	4.36e+03	94.0	n/a	312.0	n/a
S981002847	A	Uranium - ICP-Acid Digest	ug/g	94.80	<1.00e-01	1.54e+04	1.63e+04	1.58e+04	5.68	n/a	312.0	n/a
S981002847	A	Zinc - ICP-Acid Digest	ug/g	82.80	<1.00e-02	54.60	57.90	56.25	5.87	n/a	6.240	n/a
S981002847	A	Zirconium - ICP-Acid Digest	ug/g	95.80	<1.00e-02	< 6.230	< 6.660	n/a	n/a	n/a	6.240	n/a
S981002849	W	Formate by IC-Dioxex 4000/4500	ug/g	104.2	<6.20e-02	5.28e+02	432.0	480.2	20.0	n/a	261.5	n/a
S981002849	W	Fluoride by IC-Dioxex 4000/4500	ug/g	109.0	<1.20e-02	2.00e+03	1.747.0	1.37e+03	91.2	n/a	243.3	n/a
S981002849	W	Chloride by IC-Dioxex 4000/4500	ug/g	103.6	<1.70e-02	1.49e+03	1.39e+03	1.44e+03	6.94	n/a	344.7	n/a
S981002849	W	Nitrite by IC-Dioxex 4000/4500	ug/g	89.16	<1.00e-02	2.02e+04	1.76e+04	1.89e+04	13.8	n/a	2.19e+03	n/a
S981002849	W	Bromide by Ion Chromatograph	ug/g	96.76	<1.25e-01	<2.34e+03	<2.50e3	n/a	n/a	n/a	2.34e+03	n/a
S981002849	W	Nitrate by IC-Dioxex 4000/4500	ug/g	107.1	<1.35e-01	2.95e+05	3.44e+05	3.19e+05	16.0	n/a	2.82e+03	n/a
S981002849	W	Phosphate by IC-Dioxex 4000/4500	ug/g	98.57	<1.20e-01	1.74e+03	5.22e+03	1.13e+04	108	n/a	2.43e+03	n/a
S981002849	W	Sulfate by IC-Dioxex 4000/4500	ug/g	98.57	<1.35e-01	3.40e+03	<2.76e3	n/a	n/a	n/a	2.80e+03	n/a
S981002849	W	Oxalate by IC-Dioxex 4000/4500	ug/g	108.3	<1.05e-01	4.81e+03	4.00e+03	4.41e+03	18.4	n/a	2.15e+03	n/a

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002813		DSC Exotherm Dry, Calculated	Joules/g Dry	n/a	n/a	5.350	3.010	4.283	58.9	n/a	n/a	n/a
S981002813		DSC Exotherm on Perkin Elmer	Joules/g	98.91	n/a	2.655	1.457	2.056	38.8	n/a	n/a	n/a
S981002813		Alumina by ISE-Std Additions	ug/ml	113.8	<5.000	2.98e+02	1.0510	300.5	3.00	n/a	53.00	n/a
S981002813		Specific Gravity	Sp.G.	99.06	n/a	1.453	1.376	1.414	3.44	n/a	1.00e-03	n/a
S981002813		% Water by Ika on Perkin Elmer	%	97.76	<1.26e-02	32.14	52.36	52.21	0.27	n/a	n/a	n/a
S981002813		Alpha in Liquid Samples	ug/ml	117.6	<1.26e-02	1.14e+02	2.79E+3	n/a	n/a	n/a	1.90e-02	5.00e-02
S981002813		Strontium-90/90 in High Level	ug/ml	105.7	<1.20e-02	2.40e+02	2.41e+02	2.41e+02	0.42	n/a	2.00e-03	1.32e+01
S981002821	D	Al-235/240 by TRU SPEC Resin	ug/ml	97.03	<5.30e-04	4.30e+04	4.10E+02	n/a	n/a	n/a	3.50e-04	1.00e-02
S981002821	D	Aluminum ICP-Acid Dtl.	ug/ml	102.2	<5.00e-02	4.89e+04	4.33e+04	4.64e+04	0.63	n/a	30.10	n/a
S981002821	D	Boron-IC-Acid Dtl.	ug/ml	95.46	<5.00e-02	< 3.10	<3.10	33.10	0.00	n/a	30.10	n/a
S981002821	D	Barium-IC-Acid Dtl.	ug/ml	99.40	<5.00e-02	< 3.10	<3.10	n/a	n/a	n/a	30.10	n/a
S981002821	D	Bismuth-IC-Acid Dtl.	ug/ml	101.4	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S981002821	D	Calcium-IC-Acid Dtl.	ug/ml	98.80	<1.00e-02	1.16e+03	6.01e1	1.14e+03	0.78	n/a	60.10	n/a
S981002821	D	Chromium-IC-Acid Dtl.	ug/ml	95.40	<5.00e-02	5.17e+03	50.85	50.85	0.50	n/a	60.10	n/a
S981002821	D	Cobalt-IC-Acid Dtl.	ug/ml	95.80	<5.00e-02	4.65e+03	4.203	4.54e+03	5.07	n/a	300.0	n/a
S981002821	D	Pressium-IC-Acid Dtl.	ug/ml	101.6	<1.00e-02	< 2.010	<2.010	n/a	n/a	n/a	6.010	n/a
S981002821	D	Manganese-IC-Acid Dtl.	ug/ml	97.40	<1.00e-02	< 2.010	<2.010	n/a	n/a	n/a	6.010	n/a
S981002821	D	Sodium-IC-Acid Dtl.	ug/ml	100.8	<1.00e-02	2.29e+05	2.16e05	2.18e+05	1.83	n/a	60.10	n/a
S981002821	D	Silicon-IC-Acid Dtl.	ug/ml	98.40	<1.00e-02	< 13.00	<13.00	n/a	n/a	n/a	12.00	n/a
S981002821	D	Silver-IC-Acid Dtl.	ug/ml	102.6	<5.00e-02	2.02e+03	<3.01e2	200.0	2.00	n/a	36.00	n/a
S981002821	D	Zinc-IC-Acid Dtl.	ug/ml	97.40	<1.00e-02	<3.01e2	<3.01e2	n/a	n/a	n/a	30.10	n/a
S981002821	D	Zirconium-IC-Acid Dtl.	ug/ml	94.80	<1.00e-02	< 2.010	<2.010	n/a	n/a	n/a	6.010	n/a
S981002821	D	Zirconium-IC-Acid Dtl.	ug/ml	94.80	<1.00e-02	< 2.010	<2.010	n/a	n/a	n/a	6.010	n/a
S981002821	D	Zirconium-IC-Acid Dtl.	ug/ml	100.0	<1.20e-02	9.54e+02	979.0	867.4	2.58	n/a	37.17	n/a
S981002821	D	Formate by IC-Dioxex 4000/4500	ug/ml	101.1	<1.20e-02	< 51.34	<4.18e3	n/a	n/a	n/a	41.84	n/a
S981002821	D	Fluoride by IC-Dioxex 4000/4500	ug/ml	105.2	<1.70e-02	5.17e+03	5.17e03	5.23e+03	2.89	n/a	89.57	n/a
S981002821	D	Chloride by IC-Dioxex 4000/4500	ug/ml	102.5	<1.08e-01	7.95e+04	8.11e04	8.00e+04	2.73	n/a	566.3	n/a
S981002821	D	Nitrite by Ion Chromatograph	ug/ml	99.52	<1.25e-01	<6.44e+02	<6.44e2	n/a	n/a	n/a	643.3	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count	Err%
S981002821		Nitrate by IC-Dionex 4000/4500	ug/mL	99.57	<1.59e-01	1.76e-03	1.74e-03	1.75e-03	1.14	n/a	716.0	n/a
S981002821		Phosphate-IC-Dionex 4000/4500	ug/mL	104.0	<1.120e-01	1.43e-03	3.67e-03	2.56e-03	86.77	n/a	618.1	n/a
S981002821		Sulfate by IC-Dionex 4000/4500	ug/mL	101.4	<1.38e-01	4.19e-03	4.02e-03	4.11e-03	4.14	n/a	710.8	n/a
S981002821		Oxalate-IC-Dionex 4000/4500	ug/mL	106.8	<1.05e-01	5.41e-02	<5.41e-02	n/a	n/a	n/a	540.5	n/a
S981002821		Cesium-137 by GEA	uCi/mL	99.69	<1.52e-03	2.94e-02	2.97.0	295.0	1.02	n/a	n/a	0.180
S981002829		Chromium (VI) by Spec.	ug/mL	100.0	<2.58e-01	1.53e-03	1.32e-03	1.32e-03	0.75	n/a	1.00e-02	n/a
S981002829	0	1.129 by ICP/MS Acid Addition	ug/mL	104.7	<1.120e-04	<4.860	44.8660	n/a	n/a	n/a	4.86e-02	n/a
S981002829		OH- by Pot. Titration	ug/mL	100.0	n/a	5.10e-04	6.10e-04	6.10e-04	10.2	n/a	2.50e-03	n/a
S981002829		TOC by Persulfate/Coulometry	ug/mL	94.67	8.700	2.53e-03	2.64e-03	2.58e-03	4.26	n/a	40.00	n/a
S981002829	0	Uranium-235 by ICP/MS Acid Add	ug/mL	n/a	<1.120e-03	44.86e-01	44.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002829	0	Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<1.120e-03	44.86e-01	44.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002829	0	Uranium-235 by ICP/MS Acid Add	ug/mL	107.7	<1.120e-03	44.86e-01	44.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002829	0	Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<1.150e-03	45.86e-01	45.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002829	0	Uranium-238 by ICP/MS Acid Add	ug/mL	112.0	<1.120e-03	44.86e-01	44.86e-01	n/a	n/a	n/a	4.86e-01	n/a
S981002829	0	IC 99 by ICP/MS Acid Addition	ug/mL	92.68	<1.120e-03	<3.763	5.700	5.722	0.70	n/a	4.10e-02	n/a
S981002829	0	Sn 126 by ICP/MS Acid Addition	ug/mL	104	<1.50e-04	<3.043	<3.0560	n/a	n/a	n/a	3.045	n/a
S981002829	A	Neptunium-237 by ICP/MS	ug/mL	104.2	<3.47e-06	4.89e-01	<1.89e-01	n/a	n/a	n/a	1.89e-01	n/a
S981002829	A	Plutonium-239 by ICP/MS	ug/mL	104.2	<3.47e-06	4.41e-01	<1.41e-01	n/a	n/a	n/a	1.41e-01	n/a
S981002829	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<3.55e-06	<1.44e-01	<1.44e-01	n/a	n/a	n/a	1.44e-01	n/a
S981002829	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<2.76e-06	<1.13e-01	<1.13e-01	n/a	n/a	n/a	1.13e-01	n/a
S981002829	A	Pa/Am-241 by ICP/MS	ug/mL	109.7	<5.71e-06	<1.50e-01	<1.50e-01	n/a	n/a	n/a	1.50e-01	n/a

26-mar-1999 07:09:24
A-0002-1

Table 1. Data Summary Report.
BT-105 (2)

CORE NUMBER: 246R
SEGMENT #: Composite

SEGMENT PORTION: Core Composite										
Sample#	R/A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit Count Err%
S98T003383		DSC Exotherm Dry Calculated	Joules/g Dry	N/A	N/A	0.00±0.0	0.00±0.0	0.00±0.0	N/A	N/A
S98T003383		DSC Exotherm on Perkin Elmer	Joules/g	95.78	N/A	0.00±0.0	0.00±0.0	0.00±0.0	N/A	N/A
S98T003383		Ammonia by ISE-STD Additions	ug/mL	81.73	<5.00	1.03±0.03	1.01±0.03	1.02±0.03	102.0	N/A
S98T003383		% Water by TGA on Perkin Elmer	%	98.99	N/A	52.12	52.68	52.40	1.07	N/A
S98T003383		Alpha In Liquid Samples	UCI/mL	93.87	<7.90e-03	<1.39e-02	<7.90E-3	N/A	94.57	1.90e-02 5.00E+02

Table 1. Data Summary Report.
8Y-105 (2)

CORE NUMBER: 2649

SEGMENT #: Core Composite

SEGMENT POSITION: Core Composite

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	% Spk Rec.	Det Limit	Count Error
S98T0033364		Specific Gravity	Sp G.	99.14	n/a	1.259	1.404	1.331	n/a	n/a	1.00e-03	n/a
S98T0033364		Strontium-89/90 High Level	uCi/mL	100.0	<1.54e-03	1.26e-01	1.32e-01	1.29e-01	4.65	n/a	2.00e-03	5.92e+00
S98T0033364		Pu-239/240 by TRU-SPEC Resin	uCi/mL	111.6	<1.93e-05	<1.30e-04	3.46e-04	n/a	n/a	n/a	1.30e-04	9.34e+00
S98T0033364	D	Aluminum-ICP-Acid Dil.	ug/mL	106.2	<5.00e-02	3.83e+04	3.91e+04	3.87e+04	2.07	52.00	30.10	n/a
S98T0033364	D	Boron-ICP-Acid Dil.	ug/mL	106.2	<5.00e-02	38.30	39.20	38.75	2.32	102.00	30.10	n/a
S98T0033364	D	Barium-ICP-Acid Dil.	ug/mL	100.0	<5.00e-02	<3.01e-01	<3.01e-01	n/a	n/a	91.30	30.10	n/a
S98T0033364	D	Bismuth-ICP-Acid Dil.	ug/mL	101.6	<1.00e-01	<6.01e-01	<6.01e-01	n/a	n/a	100.0	60.10	n/a
S98T0033364	D	Calcium-ICP-Acid Dil.	ug/mL	99.00	<1.00e-02	1.52e+03	1.54e+03	1.53e+03	1.31	100.0	60.10	n/a
S98T0033364	D	Iron-ICP-Acid Dil.	ug/mL	95.60	<5.00e-02	<3.01e-03	<3.01e-03	n/a	n/a	97.10	30.10	n/a
S98T0033364	D	Potassium-ICP-Acid Dil.	ug/mL	97.20	<5.00e-01	4.51e+03	4.50e+03	4.50e+03	0.22	81.60	300.0	n/a
S98T0033364	D	Lithium-ICP-Acid Dil.	ug/mL	100.8	<1.00e-02	<6.01e-01	<6.01e-01	n/a	n/a	94.00	60.10	n/a
S98T0033364	D	Manganese-ICP-Acid Dil.	ug/mL	97.80	<1.00e-02	<6.01e-01	<6.01e-01	n/a	n/a	93.10	60.10	n/a
S98T0033364	D	Sodium-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	2.14e+05	2.20e+05	2.17e+05	2.76	-5.91e2	60.10	n/a
S98T0033364	D	Nickel-ICP-Acid Dil.	ug/mL	98.40	<2.00e-02	<12.00	<1.20e1	n/a	n/a	96.40	12.00	n/a
S98T0033364	D	Silicon-ICP-Acid Dil.	ug/mL	108.2	<5.00e-02	2.75e+02	282.0	278.5	2.51	111.0	30.10	n/a
S98T0033364	D	Strontium-ICP-Acid Dil.	ug/mL	98.80	<5.00e-01	<3.01e+02	<3.01e2	n/a	n/a	95.00	300.0	n/a
S98T0033364	D	Zinc-ICP-Acid Dil.	ug/mL	95.40	<1.00e-02	<6.01e-01	<6.01e0	n/a	n/a	95.90	60.10	n/a
S98T0033364	D	Zirconium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	<6.01e-01	<6.01e0	n/a	n/a	99.30	60.10	n/a
S98T0033364		Formate by Ion Chromatograph	ug/mL	94.96	<2.10e-02	9.80e+02	1.02e+03	999.8	4.00	98.32	319.4	n/a
S98T0033364		Fluoride-IC-Bioxon 4000/4500	ug/mL	110.7	<1.20e-02	3.06e+02	3.06e+02	305.9	0.00	117.1	61.81	n/a
S98T0033364		Chloride-IC-Bioxon 4000/4500	ug/mL	100.5	<1.70e-02	4.36e+03	4.28e+03	4.32e+03	1.85	100.0	87.57	n/a
S98T0033364		Nitrite-IC - Bioxon 4000/4500	ug/mL	98.33	<1.60e-01	6.76e+02	6.95e+02	685.5	2.77	87.03	643.9	n/a
S98T0033364		Bromide by Ion Chromatograph	ug/mL	96.25	<1.25e-01	1.82e+05	1.81e+05	1.82e+05	0.55	86.99	716.0	n/a
S98T0033364		Nitrate by IC-Bioxon 4000/4500	ug/mL	96.79	<1.39e-01	1.05e+03	957.0	1.00e+03	9.27	88.48	618.1	n/a
S98T0033364		Phosphate-IC-Bioxon 4000/4500	ug/mL	93.97	<1.20e-01	2.55e+03	2.60e+03	2.58e+03	1.94	93.42	710.8	n/a
S98T0033364		Sulfate by IC-Bioxon 4000/4500	ug/mL	96.55	<1.05e-01	<5.41e+02	590.0	n/a	n/a	97.74	540.9	n/a
S98T0033364		Oxalate-IC-Bioxon 4000/4500	ug/mL	101.4	<1.65e-03	1.96e+02	201.0	198.5	2.52	n/a	n/a	0.120
S98T0033364		Cesium-137 by GEA	uCi/mL	100.4	<1.00e-03	1.53e+03	1.55e+03	1.54e+03	1.30	98.20	1.00e+02	n/a
S98T0033365	D	Chromium (VI) by Spec.	ug/mL	99.80	<2.58e-01	<4.86e-01	<4.86e0	n/a	n/a	n/a	4.86e0	n/a
S98T0033365	D	1129 by ICP/MS Acid Addition	ug/mL	105.0	<1.20e-04	<4.86e-01	<4.86e-01	4.60e+04	1.96	n/a	2.50e+03	n/a
S98T0033365		Oil - by Pot. Titration	ug/mL	90.67	<2500.0	2.28e+03	2.28e+03	2.28e+03	0.00	96.00	40.00	n/a
S98T0033365		TOC by Persulfate/Coulometry	ug/mL	90.67	<2500.0	2.28e+03	2.28e+03	2.28e+03	0.00	96.00	40.00	n/a
S98T0033365	D	Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	n/a	4.86e-01	n/a
S98T0033365	D	Uranium-235 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	n/a	4.86e-01	n/a
S98T0033365	D	Uranium-236 by ICP/MS Acid Add	ug/mL	100.0	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	94.94	4.86e-01	n/a
S98T0033365	D	Uranium-238 by ICP/MS Acid Add	ug/mL	n/a	<1.60e-05	<4.86e-01	<4.86e-1	n/a	n/a	n/a	4.86e-01	n/a
S98T0033365	D	Uranium-238 by ICP/MS Acid Add	ug/mL	104.0	<1.20e-05	1.284	1.160	1.222	9.84	83.56	8.10e+02	n/a
S98T0033365	D	Tc 99 by ICP/MS Acid Addition	ug/mL	85.37	<2.00e-06	<4.86e-01	<4.86e-1	4.750	4.800	93.50	8.10e+02	n/a
S98T0033365	D	Sn 126 by ICP/MS Acid Addition	ug/mL	n/a	<1.50e-04	<6.075	<6.08e0	n/a	n/a	n/a	6.075	n/a
S98T0033365	D	Thorium-229 by ICP/MS	ug/mL	n/a	<4.46e-06	<1.81e-01	<1.81e-1	n/a	n/a	n/a	1.81e-01	n/a
S98T0033365	D	Neptunium-237 by ICP/MS	ug/mL	119.6	<6.67e-06	<1.89e-01	<1.89e-1	n/a	n/a	115.9	1.89e-01	n/a
S98T0033365	D	Plutonium-239 by ICP/MS	ug/mL	106.2	<3.47e-06	<1.41e-01	<1.41e-1	n/a	n/a	99.10	1.41e-01	n/a
S98T0033365	D	Plutonium-240 by ICP/MS	ug/mL	n/a	<3.55e-06	<1.44e-01	<1.44e-1	n/a	n/a	n/a	1.44e-01	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003385	D	Plutonium-242 by ICP/MS	ug/mL	D/A	<2.78e-06	<1.13e-01	<1.13e-1	D/A	D/A	D/A	1.13e-01	D/A
S98T003385	D	Pu/Am-241 by ICP/MS	ug/mL	109.7	<3.71e-06	<1.50e-01	<1.50e-1	D/A	D/A	105.3	1.50e-01	D/A

CORE NUMBER: 250
SEGMENT #: 3

SEGMENT PORTION: L Lower Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002736		Bulk Density of Sample	g/mL	N/A	N/A	dry	N/A	N/A	N/A	N/A	5.00e-01	N/A
S98T002737		DSC Exotherm Dry, Calculated	Joules/g DRY	N/A	N/A	20.78	21.83	22.30	13.7	N/A	N/A	N/A
S98T002738		DSC Exotherm on Perkin Elmer	Joules/g	100.6	N/A	18.19	20.86	19.52	13.7	N/A	N/A	N/A
S98T002739		% Water by IGA on Perkin Elmer		98.99	N/A	11.71	13.22	12.46	12.1	N/A	N/A	N/A
S98T002740		TIC by Acid/Coulometry	ug/g	99.50	4.500	2.40e+04	1.46e+04	1.93e+04	48.7	24.30	5.000	N/A
S98T002741		TOC by Persulfate/Coulometry	ug/g	94.00	2.100	3.57e+03	3.16e+03	3.36e+03	11.6	91.70	40.00	N/A
S98T002742	F	Strontium-89/90 High Level	uCi/g	97.11	6.70e-02	3.330	3.360	3.85	5.34	N/A	2.00e+03	1.06E+00
S98T002743	F	Aluminum - ICP-Fusion	ug/g	89.20	4.50e-02	5.27e+02	5.22e+04	5.24e+04	0.95	75.60	1.02e+03	N/A
S98T002744	F	Boron - ICP-Fusion	ug/g	100.6	4.50e-02	1.02e+03	1.02e3	N/A	N/A	92.10	1.02e+03	N/A
S98T002745	F	Barium - ICP-Fusion	ug/g	99.00	4.50e-02	1.02e+03	1.02e3	N/A	N/A	86.50	1.02e+03	N/A
S98T002746	F	Bismuth - ICP-Fusion	ug/g	101.41	1.00e-01	1.02e+03	1.02e3	N/A	N/A	93.00	2.05e+03	N/A
S98T002747	F	Calcium - ICP-Fusion	ug/g	104.01	1.00e-01	1.02e+03	1.02e3	N/A	N/A	98.00	2.05e+03	N/A
S98T002748	F	Chromium - ICP-Fusion	ug/g	101.21	1.00e-02	3.89e+02	412.0	400.5	5.74	95.20	205.0	N/A
S98T002749	F	Iron - ICP-Fusion	ug/g	97.80	4.50e-02	1.02e+03	1.02e3	N/A	N/A	92.40	1.02e+03	N/A
S98T002750	F	Lithium - ICP-Fusion	ug/g	103.21	1.00e-02	1.02e+03	1.02e3	N/A	N/A	87.90	205.0	N/A
S98T002751	F	Manganese - ICP-Fusion	ug/g	100.01	1.00e-02	1.02e+03	1.02e3	N/A	N/A	91.30	205.0	N/A
S98T002752	F	Sodium - ICP-Fusion	ug/g	102.61	1.00e-01	2.15e+03	2.16e+03	2.16e+03	0.46	26.60	2.05e+03	N/A
S98T002753	F	Nickel - ICP-Fusion	ug/g	100.21	1.65e-01	4.69e+03	5.22e+03	5.22e+03	20.5	93.40	409.0	N/A
S98T002754	F	Silicon - ICP-Fusion	ug/g	104.21	4.50e-02	2.11e+03	2.05e+03	2.08e+03	2.88	99.40	1.02e+03	N/A
S98T002755	F	Uranium - ICP-Fusion	ug/g	98.90	4.50e-01	1.02e+04	1.02e4	N/A	N/A	89.50	1.02e+04	N/A
S98T002756	F	Uranium - ICP-Fusion	ug/g	95.40	1.00e-02	1.02e+02	1.02e+02	N/A	N/A	91.90	205.0	N/A
S98T002757	F	Zinc - ICP-Fusion	ug/g	98.60	1.00e-02	1.02e+02	1.02e+02	N/A	N/A	89.90	205.0	N/A
S98T002758	F	Zirconium - ICP-Fusion	ug/g	102.31	1.87e-01	27.63	28.80	28.21	4.26	N/A	N/A	0.610
S98T002759	F	Alpha of Digested Solid	uCi/g	89.44	2.45e-03	1.240e-01	1.240e-01	N/A	N/A	85.93	3.35e-01	5.00E-02
S98T002760	A	1199 by ICP/MS ActiGLA505159	ug/g	109.01	1.20e-03	4.92e-01	4.68e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002741	A	Uranium-233 by ICP/MS ActiGL159	ug/g	N/A	1.32e-04	4.92e-01	4.68e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002762	A	Uranium-235 by ICP/MS ActiGL159	ug/g	N/A	1.32e-04	4.92e-01	4.68e-01	N/A	N/A	N/A	4.92e-01	N/A
S98T002743	A	Uranium-235 by ICP/MS ActiGL159	ug/g	98.59	1.32e-04	1.330	1.510	1.420	12.7	90.72	4.92e-01	N/A
S98T002764	A	Uranium-238 by ICP/MS ActiGL159	ug/g	N/A	1.76e-04	6.57e-01	6.24e-01	N/A	N/A	N/A	6.57e-01	N/A
S98T002745	A	Uranium-238 by ICP/MS ActiGL159	ug/g	94.50	1.32e-04	1.88e+02	218.0	202.8	14.8	88.50	4.92e-01	N/A
S98T002766	A	T999 by ICP/MS ActiGLA505159	ug/g	92.19	5.39e-05	7.10e-01	7.99e-01	7.54e-01	11.8	90.87	8.20e-02	N/A
S98T002747	A	Sm126 ICP/MS ActiGLA505159	ug/g	N/A	1.65e-03	6.15e	5.85e0	N/A	N/A	N/A	6.15	N/A
S98T002748	A	Neptunium-237 by ICP/MS	ug/g	118.9	4.231e-05	4.61e-02	4.81e-02	N/A	N/A	N/A	8.60e-02	N/A
S98T002749	A	Plutonium-239 by ICP/MS	ug/g	99.60	4.410e-05	1.510e-01	1.42e-01	N/A	N/A	N/A	1.50e-01	N/A
S98T002750	A	Plutonium-240 by ICP/MS	ug/g	99.60	1.12e-05	4.19e-02	3.96e-02	N/A	N/A	N/A	4.20e-02	N/A
S98T002751	A	Plutonium-242 by ICP/MS	ug/g	N/A	3.15e-05	1.20e-02	1.22e-02	N/A	N/A	N/A	1.29e-01	N/A
S98T002742	A	Pu/Am-241 by ICP/MS	ug/g	103.9	1.185e-05	4.88e-02	5.34e-02	N/A	N/A	N/A	6.90e-02	N/A
S98T002743	A	Aluminum - ICP-Acid Digest	ug/g	96.40	1.339e-01	3.17e+04	5.70e+04	5.44e+04	9.75	210.0	51.00	N/A
S98T002744	A	Boron - ICP-Acid Digest	ug/g	107.4	2.27e-01	1.35e+02	60.80	97.90	75.8	100.4	51.00	N/A
S98T002745	A	Barium - ICP-Acid Digest	ug/g	93.00	4.50e-02	31.00	44.8e1	N/A	N/A	94.00	51.00	N/A
S98T002746	A	Bismuth - ICP-Acid Digest	ug/g	91.80	1.00e-01	1.02e+02	49.70e1	N/A	N/A	98.20	102.0	N/A
S98T002747	A	Calcium - ICP-Acid Digest	ug/g	103.01	1.57e-01	1.47e+02	169.0	158.0	13.9	106.2	102.0	N/A
S98T002748	A	Chromium - ICP-Acid Digest	ug/g	94.20	1.00e-02	3.64e+02	412.0	388.0	12.4	102.4	10.20	N/A
S98T002749	A	Iron - ICP-Acid Digest	ug/g	91.20	49.00e-02	6.05e+02	478.0	541.5	23.5	79.40	51.00	N/A

26-mar--1999 07:13:42
A-0002-1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002740	A	Potassium -ICP-Acid Digest	ug/g	96.20	< 5.00e-01	7.60e+02	970.0	865.0	24.3	74.80	510.0	n/a
S98T002740	A	Lithium -ICP-Acid Digest	ug/g	97.60	< 1.00e-02	< 10.20	< 9.70e0	n/a	n/a	96.20	10.20	n/a
S98T002740	A	Manganese -ICP-Acid Digest	ug/g	97.20	< 1.00e-02	22.00	24.50	23.25	10.8	96.60	10.20	n/a
S98T002740	A	Sodium -ICP-Acid Digest	ug/g	139.2	1.150	2.15e+05	2.23e+05	2.19e+05	3.65	1.49e+03	102.0	n/a
S98T002740	A	Nickel -ICP-Acid Digest	ug/g	91.86	< 2.00e-02	69.00	73.50	71.25	6.32	99.60	20.40	n/a
S98T002740	A	Silicon -ICP-Acid Digest	ug/g	85.86	2.130	5.20e+02	517.0	518.5	0.58	126.2	51.00	n/a
S98T002740	A	Uranium -ICP-Acid Digest	ug/g	93.60	< 5.00e-01	< 5.10e+02	< 4.85e2	n/a	n/a	104.0	510.0	n/a
S98T002740	A	Zinc -ICP-Acid Digest	ug/g	93.60	< 1.00e-02	< 10.20	< 9.70e0	n/a	n/a	96.00	10.20	n/a
S98T002740	A	Zirconium -ICP-Acid Digest	ug/g	93.00	< 1.00e-02	< 10.20	< 9.70e0	n/a	n/a	89.80	10.20	n/a
S98T002740	W	Formate by IC-Dionex 4000/4500	ug/g	102.2	< 6.20e-02	< 2.61e+02	< 2.63e2	n/a	n/a	97.66	260.8	n/a
S98T002740	W	Fluoride-IC-Dionex 4000/4500	ug/g	97.74	< 1.20e-02	5.77e+02	2.43e+03	2.54e+03	8.66	109.8	122.6	n/a
S98T002740	W	Chloride-IC-Dionex 4000/4500	ug/g	89.23	< 1.20e-01	7.19e+02	8.00e+03	601.5	8.15	96.11	173.7	n/a
S98T002740	W	Nitrate by Ion chromatograph	ug/g	92.32	< 1.25e-01	< 1.26e+03	< 1.29e3	n/a	n/a	88.04	1.10e+03	n/a
S98T002740	W	Nitrate by IC-Dionex 4000/4500	ug/g	100.2	< 1.30e-01	3.63e+05	3.39e+05	3.51e+05	6.84	94.03	1.28e+03	n/a
S98T002740	W	Phosphate by IC-Dionex 4000/4500	ug/g	94.89	< 1.20e-01	2.01e+04	1.63e+04	1.82e+04	20.9	100.5	1.23e+03	n/a
S98T002740	W	Sulfate by IC-Dionex 4000/4500	ug/g	93.75	< 1.38e-01	6.46e+03	5.18e+03	4.91e+03	10.2	96.24	1.41e+03	n/a
S98T002740	W	Oxalate by IC-Dionex 4000/4500	ug/g	102.4	< 1.09e-01	8.56e+03	9.18e+03	8.87e+03	6.99	105.6	1.07e+03	n/a

Table 1. Data Summary Report.
BY-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count	Err%
S981002905		Bulk Density of Sample	g/mL	n/a	n/a	n/a	n/a	n/a	n/a	5.00e-01	n/a	n/a
S981002910		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002910		DSC Exotherm on Perkin Elmer	Joules/g	98.52	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002910		% Water by TGA on Perkin Elmer	%	99.26	n/a	5.360	10.21	7.785	62.3	n/a	n/a	n/a
S981002910		TIC by Acid/Coulometry	ug/g	95.85	5.900	1.90e+03	888.0	1.39e+03	72.6	n/a	5.000	n/a
S981002910		TOC by Persulfate/Coulometry	ug/g	95.67	5.200	1.69e+03	1.58e+03	1.64e+03	6.73	n/a	40.00	n/a
S981002918		F Strontium-89/90 High Level	uCi/g	107.7	<1.10e-01	1.950	1.610	1.780	19.1	n/a	9.00e-03	2.14e+03
S981002918		F Boron - ICP-Fusion	ug/g	96.40	<5.00e-02	4.01e+04	2.77e+04	3.39e+04	36.6	n/a	1.02e+03	n/a
S981002918		F Bismuth - ICP-Fusion	ug/g	103.0	<5.00e-02	<1.02e+03	<1.01e3	n/a	n/a	n/a	1.02e+03	n/a
S981002918		F Barium - ICP-Fusion	ug/g	96.60	<5.00e-02	<1.02e+03	<2.01e3	n/a	n/a	n/a	2.04e+03	n/a
S981002918		F Calcium - ICP-Fusion	ug/g	103.0	<1.00e-01	<2.04e+03	<2.01e3	n/a	n/a	n/a	2.04e+03	n/a
S981002918		F Chromium - ICP-Fusion	ug/g	99.60	<1.00e-02	6.10e+02	382.0	496.0	46.0	n/a	204.0	n/a
S981002918		F Iron - ICP-Fusion	ug/g	95.60	<5.00e-02	<1.02e+03	<1.01e3	n/a	n/a	n/a	1.02e+03	n/a
S981002918		F Lithium - ICP-Fusion	ug/g	99.60	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	n/a	204.0	n/a
S981002918		F Manganese - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	n/a	204.0	n/a
S981002918		F Nickel - ICP-Fusion	ug/g	98.20	<1.00e-01	2.11e+05	2.27e+05	2.19e+05	7.31	n/a	2.04e+03	n/a
S981002918		F Radium - ICP-Fusion	ug/g	99.80	1.27e-01	4.03e+03	3.55e+03	3.79e+03	12.7	n/a	408.0	n/a
S981002918		F Strontium - ICP-Fusion	ug/g	104.0	<5.00e-02	<1.04e+03	1.40e+03	1.52e+03	15.8	n/a	1.02e+03	n/a
S981002918		F Uranium - ICP-Fusion	ug/g	96.70	<5.00e-01	<2.04e+02	<1.01e4	n/a	n/a	n/a	1.02e+04	n/a
S981002918		F Zinc - ICP-Fusion	ug/g	95.00	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	n/a	204.0	n/a
S981002918		F Zirconium - ICP-Fusion	ug/g	97.20	<1.00e-02	<2.04e+02	<2.01e2	n/a	n/a	n/a	204.0	n/a
S981002918		F Alpha of Digestion Solid	uCi/g	103.4	<1.75e-01	26.34	21.70	24.02	19.2	n/a	n/a	0.770
S981002918		F Alpha of Digestion Solid	uCi/g	95.56	<3.27e-01	<2.29e-01	<2.60E-01	n/a	n/a	n/a	3.98e-01	5.00e-02
S981002926		A Uranium-235 by ICP/MS Act159	ug/g	109.0	<1.20e-03	<4.77e	<4.83e0	n/a	n/a	n/a	4.77e	n/a
S981002926		A Uranium-235 by ICP/MS Act159	ug/g	n/a	<1.20e-04	44.78e-01	44.83e-1	n/a	n/a	n/a	4.78e-01	n/a
S981002926		A Uranium-235 by ICP/MS Act159	ug/g	105.6	<1.20e-04	9.87e-01	9.87e-01	8.60e-01	29.5	n/a	4.78e-01	n/a
S981002926		A Uranium-235 by ICP/MS Act159	ug/g	n/a	<1.20e-04	<6.37e-01	<6.44e-1	n/a	n/a	n/a	6.37e-01	n/a
S981002926		A Uranium-238 by ICP/MS Act159	ug/g	106.0	<4.33e-04	1.39e+02	113.0	125.8	20.6	n/a	4.78e-01	n/a
S981002926		A Uranium-238 by ICP/MS Act159	ug/g	91.22	<3.64e-04	8.44e-01	7.33e-01	7.88e-01	14.1	n/a	8.00e-02	n/a
S981002926		A Sm126 by ICP/MS Act160 LA-305159	ug/g	n/a	<1.65e-03	<5.970	<6.04e0	n/a	n/a	n/a	5.970	n/a
S981002926		A Neptunium-237 by ICP/MS	ug/g	116.4	<1.63e-03	<5.88e-02	<5.95e-2	n/a	n/a	n/a	5.90e-02	n/a
S981002926		A Plutonium-239 by ICP/MS	ug/g	98.51	<3.54e-03	<1.29e-01	<1.30e-1	n/a	n/a	n/a	1.29e-01	n/a
S981002926		A Plutonium-239 by ICP/MS	ug/g	n/a	<2.40e-03	<8.67e-02	<8.7e-2	n/a	n/a	n/a	8.70e-02	n/a
S981002926		A Plutonium-240 by ICP/MS	ug/g	n/a	<1.69e-02	<6.12e-02	<6.19e-2	n/a	n/a	n/a	6.10e-02	n/a
S981002926		A Plutonium-242 by ICP/MS	ug/g	102.3	<1.97e-02	<7.21e-02	<7.21e-2	n/a	n/a	n/a	7.10e-02	n/a
S981002926		A Pu/Am-241 by ICP/MS	ug/g	94.80	9.70e-02	4.09e+04	3.50e+04	3.80e+04	15.5	n/a	49.50	n/a
S981002926		A Lanthanum - ICP-Acid Digest	ug/g	104.0	<2.29e-01	1.30e+02	106.0	118.0	20.3	n/a	49.50	n/a
S981002926		A Boron - ICP-Acid Digest	ug/g	94.80	<5.00e-02	<3.49e1	<5.01e1	n/a	n/a	n/a	49.50	n/a
S981002926		A Barium - ICP-Acid Digest	ug/g	91.40	<1.00e-01	<99.00	<1.00e2	n/a	n/a	n/a	99.00	n/a
S981002926		A Bismuth - ICP-Acid Digest	ug/g	97.20	1.192e-01	1.62e+02	160.0	161.0	1.24	n/a	99.00	n/a
S981002926		A Calcium - ICP-Acid Digest	ug/g	93.60	<1.00e-02	6.31e+02	561.0	606.0	1.49	n/a	9.900	n/a
S981002926		A Chromium - ICP-Acid Digest	ug/g	90.60	<5.00e-02	2.80e+02	249.0	264.5	11.7	n/a	49.50	n/a
S981002926		A Iron - ICP-Acid Digest	ug/g									

124

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002926	A	Potassium - ICP-Acid Digest	ug/g	85.60	< 5.00e-01	6.53e+02	< 5.01e2	n/a	n/a	n/a	930.0	n/a
S981002926	A	Lithium - ICP-Acid Digest	ug/g	99.00	< 1.00e-02	< 9.900	< 1.00e1	n/a	n/a	n/a	9.900	n/a
S981002926	A	Manganese - ICP-Acid Digest	ug/g	91.20	< 1.00e-02	13.60	11.20	12.40	19.2	n/a	9.900	n/a
S981002926	A	Sodium - ICP-Acid Digest	ug/g	115.0	8.58e-01	2.20e+05	2.11e+05	2.16e+05	4.18	n/a	99.00	n/a
S981002926	A	Nickel - ICP-Acid Digest	ug/g	90.40	< 1.00e-02	88.50	85.60	87.05	3.33	n/a	19.80	n/a
S981002926	A	Silicon - ICP-Acid Digest	ug/g	154.2	6.93e-01	5.54e+02	455.0	504.5	19.2	n/a	49.50	n/a
S981002926	A	Uranium - ICP-Acid Digest	ug/g	94.80	< 1.00e-01	44.95e+02	< 5.01e2	n/a	n/a	n/a	495.0	n/a
S981002926	A	Zinc - ICP-Acid Digest	ug/g	82.80	< 1.00e-02	< 9.900	< 1.00e1	n/a	n/a	n/a	9.900	n/a
S981002926	A	Formate by IC-Dionex 4000/4500	ug/g	93.80	< 1.00e-02	< 9.900	< 1.00e1	n/a	n/a	n/a	9.900	n/a
S981002931	M	Fluoride-IC-Dionex 4000/4500	ug/g	91.41	< 5.20e-02	< 2.59e+02	< 2.53e2	n/a	n/a	93.42	258.7	n/a
S981002931	M	Chloride-IC-Dionex 4000/4500	ug/g	108.3	< 1.70e-02	7.23e+02	806.0	744.4	10.9	n/a	300.9	n/a
S981002931	M	Nitrite-IC - Dionex 4000/4500	ug/g	89.16	< 1.08e-01	8.81e+03	8.06e+03	8.44e+03	8.89	n/a	341.3	n/a
S981002931	M	Bromide by Ion Chromatograph	ug/g	96.76	< 1.25e-01	5.37e+03	< 2.16e3	n/a	n/a	n/a	2.17e+03	n/a
S981002931	M	Nitrate by IC-Dionex 4000/4500	ug/g	107.1	< 1.39e-01	5.79e+05	5.79e+05	5.58e+05	7.53	n/a	2.51e+03	n/a
S981002931	M	Phosphate-IC-Dionex 4000/4500	ug/g	99.27	< 1.20e-01	4.46e+03	4.81e+03	4.64e+03	7.55	n/a	2.79e+03	n/a
S981002931	M	Sulfate by IC-Dionex 4000/4500	ug/g	98.28	< 1.38e-01	3.39e+03	3.01e+03	3.20e+03	11.9	n/a	2.71e+03	n/a
S981002931	M	Oxalate-IC-Dionex 4000/4500	ug/g	108.7	< 1.03e-01	4.16e+03	3.48e+03	3.82e+03	17.8	n/a	2.71e+03	n/a

Table 1. Data Summary Report.
BY-105 (2)

CORE NUMBER: 250
SEGMENT #: 5

SEGMENT PORTION: L Lower Half of Segment

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002906		Bulk Density of Sample	g/mL	D/A	D/A	dry	n/a	n/a	n/a	5.00e-01	n/a
S98T002911		OSC Exotherm Dry, Calculated	Joules/g Dry	D/A	D/A	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S98T002911		OSC Exotherm on Perkin Elmer	Joules/g	98.52	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S98T002911		% Water by TGA on Perkin Elmer	%	99.26	n/a	2.990	4.010	3.500	2.35	112.0	5.000
S98T002911		TIC by Acid/Coulometry	ug/g	95.85	5.900	3.13e+03	3.13e+03	3.08e+03	2.67	92.10	40.00
S98T002911		TIC by Persulfate/Coulometry	ug/g	95.67	5.200	1.14e+03	1.14e+03	1.12e+03	2.67	92.10	40.00
S98T002919		F Strontium-89/90 High Level	uc/g	99.08	3.00e-03	1.80e-03	1.64e+03	1.72e+03	9.30	94.60	1.03e+03
S98T002919		F Aluminum - ICP-Fusion	ug/g	96.40	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	90.50	1.03e+03
S98T002919		F Boron - ICP-Fusion	ug/g	103.0	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	90.50	1.03e+03
S98T002919		F Barium - ICP-Fusion	ug/g	96.60	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	96.60	2.05e+03
S98T002919		F Bismuth - ICP-Fusion	ug/g	100.4	1.00e-01	2.05e+03	2.08e+03	n/a	n/a	105.0	2.05e+03
S98T002919		F Calcium - ICP-Fusion	ug/g	103.0	1.00e-01	2.05e+03	2.08e+03	n/a	n/a	92.90	205.0
S98T002919		F Chromium - ICP-Fusion	ug/g	99.60	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	94.50	1.03e+03
S98T002919		F Iron - ICP-Fusion	ug/g	99.60	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	92.10	205.0
S98T002919		F Lithium - ICP-Fusion	ug/g	99.60	1.00e-02	2.05e+03	2.08e+03	n/a	n/a	95.30	205.0
S98T002919		F Manganese - ICP-Fusion	ug/g	98.20	1.00e-01	2.05e+03	2.08e+03	n/a	n/a	90.60	2.05e+03
S98T002919		F Sodium - ICP-Fusion	ug/g	99.80	1.27e-01	2.05e+03	2.08e+03	n/a	n/a	98.80	410.0
S98T002919		F Nickel - ICP-Fusion	ug/g	99.80	1.27e-01	2.05e+03	2.08e+03	n/a	n/a	108.0	1.03e+03
S98T002919		F Silicon - ICP-Fusion	ug/g	104.0	5.00e-02	1.03e+03	1.04e+03	n/a	n/a	89.50	1.03e+04
S98T002919		F Uranium - ICP-Fusion	ug/g	96.70	5.00e-03	2.05e+03	2.08e+03	n/a	n/a	96.30	205.0
S98T002919		F Zinc - ICP-Fusion	ug/g	95.00	1.00e-03	2.05e+03	2.08e+03	n/a	n/a	93.80	205.0
S98T002919		F Zirconium - ICP-Fusion	ug/g	97.20	1.00e-02	2.05e+03	2.08e+03	n/a	n/a	93.80	205.0
S98T002919		F Cesium-137 by GEA	uc/g	103.4	1.70e-01	2.05e+03	2.08e+03	n/a	n/a	2.11	n/a
S98T002919		F Alpha of Digested Solid	uc/g	95.36	1.70e-01	2.05e+03	2.08e+03	n/a	n/a	84.20	4.01e+01
S98T002927		A 1129 by ICP/MS ActidigLA505159	ug/g	109.0	1.70e-03	4.726	4.600	n/a	n/a	n/a	4.726
S98T002927		A Uranium-235 by ICP/MS ActidigLA505159	ug/g	n/a	1.70e-02	4.73e-01	4.60e-1	n/a	n/a	n/a	4.73e-01
S98T002927		A Uranium-238 by ICP/MS ActidigLA505159	ug/g	n/a	1.70e-04	4.73e-01	4.60e-1	n/a	n/a	n/a	4.73e-01
S98T002927		A Uranium-235 by ICP/MS ActidigLA505159	ug/g	105.6	1.70e-04	6.85e-01	6.59e-1	6.72e-1	3.87	102.7	6.73e-01
S98T002927		A Uranium-238 by ICP/MS ActidigLA505159	ug/g	105.6	1.70e-04	6.85e-01	6.59e-1	6.72e-1	3.87	102.7	6.73e-01
S98T002927		A Uranium-235 by ICP/MS ActidigLA505159	ug/g	105.6	1.70e-04	6.85e-01	6.59e-1	6.72e-1	3.87	102.7	6.73e-01
S98T002927		A Uranium-238 by ICP/MS ActidigLA505159	ug/g	105.6	1.70e-04	6.85e-01	6.59e-1	6.72e-1	3.87	102.7	6.73e-01
S98T002927		A Tc99 by ICP/MS ActidigLA505159	ug/g	91.22	4.33e-04	97.27	99.60	98.44	2.34	103.5	4.73e-01
S98T002927		A Sn126 ICP/MS ActidigLA505159	ug/g	91.22	4.33e-04	97.27	99.60	98.44	2.34	103.5	4.73e-01
S98T002927		A Neptunium-237 by ICP/MS	ug/g	111.0	1.65e-03	5.907	5.750	n/a	n/a	113.9	5.907
S98T002927		A Neptunium-239 by ICP/MS	ug/g	98.51	5.40e-03	5.82e-02	5.64e-2	n/a	n/a	95.12	1.27e-01
S98T002927		A Plutonium-239 by ICP/MS	ug/g	98.51	5.40e-03	5.82e-02	5.64e-2	n/a	n/a	95.12	1.27e-01
S98T002927		A Plutonium-240 by ICP/MS	ug/g	102.3	1.69e-05	4.05e-02	4.89e-2	n/a	n/a	100.9	7.00e-02
S98T002927		A Plutonium-242 by ICP/MS	ug/g	102.3	1.69e-05	4.05e-02	4.89e-2	n/a	n/a	100.9	7.00e-02
S98T002927		A Pu/Am-241 by ICP/MS	ug/g	94.80	9.70e-02	4.72e+03	4.54e+03	4.63e+03	3.89	100.3	49.00
S98T002927		A Aluminum - ICP-Acid Digest	ug/g	104.0	5.29e-01	98.90	159.0	128.9	46.6	103.6	49.00
S98T002927		A Boron - ICP-Acid Digest	ug/g	94.80	5.00e-01	< 49.00	< 47.6e1	n/a	n/a	93.00	49.00
S98T002927		A Barium - ICP-Acid Digest	ug/g	91.40	1.00e-01	< 98.00	< 9.53e1	n/a	n/a	102.2	98.00
S98T002927		A Bismuth - ICP-Acid Digest	ug/g	97.20	1.92e-01	1.14e+02	132.0	123.0	14.6	102.6	98.00
S98T002927		A Calcium - ICP-Acid Digest	ug/g	93.00	1.00e-02	1.81e+02	187.0	184.0	3.26	99.60	9.800
S98T002927		A Chromium - ICP-Acid Digest	ug/g	90.60	5.00e-02	91.40	117.0	104.2	24.6	98.60	49.00

Sample#	R	#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002927	A	Potassium	-IP-Acid Digest	ug/g	85.60	<5.00e-01	4.490e+02	4.496e+02	n/a	n/a	128.8	940.0	n/a
S981002927	A	Lithium	-IP-Acid Digest	ug/g	89.00	<1.00e-02	< 9.800	< 9.560	n/a	n/a	94.60	9.800	n/a
S981002927	A	Manganese	-IP-Acid Digest	ug/g	11.20	<1.00e-02	< 9.800	< 9.560	n/a	n/a	94.60	9.800	n/a
S981002927	A	Sodium	-IP-Acid Digest	ug/g	113.0	8.25e-01	2.44e+03	2.44e+03	2.44e+03	0.0	490.0	98.00	n/a
S981002927	A	Nickel	-IP-Acid Digest	ug/g	19.42	<2.00e-02	8.7e+02	7.7e+02	7.7e+02	13.5	98.20	19.40	n/a
S981002927	A	Silicon	-IP-Acid Digest	ug/g	19.42	<2.00e-02	3.22e+03	3.32e+03	3.77e+03	29.2	177.0	49.00	n/a
S981002927	A	Vanadium	-IP-Acid Digest	ug/g	85.80	<5.00e-01	4.490e+02	4.496e+02	n/a	n/a	97.20	940.0	n/a
S981002927	A	Zinc	-IP-Acid Digest	ug/g	85.80	<5.00e-01	4.490e+02	4.496e+02	n/a	n/a	97.20	940.0	n/a
S981002927	A	Zirconium	-IP-Acid Digest	ug/g	92.53	<1.00e-02	< 9.800	< 9.560	n/a	n/a	96.20	9.800	n/a
S981002932	W	Formate by IC-Dionex	4000/4500	ug/g	108.3	<4.20e-02	<2.60e+03	<2.44e+03	n/a	n/a	96.88	260.2	n/a
S981002932	W	Fluoride by IC-Dionex	4000/4500	ug/g	108.3	<4.20e-02	1.72e+03	1.77e+03	1.74e+03	2.83	107.4	262.2	n/a
S981002932	W	Chloride by IC-Dionex	4000/4500	ug/g	103.4	<1.70e-02	4.27e+02	<5.48e+02	n/a	n/a	94.97	363.1	n/a
S981002932	W	Nitrate by IC-Dionex	4000/4500	ug/g	89.16	<1.09e-01	4.10e+03	<2.56e+03	4.19e+03	4.53	85.05	2.18e+03	n/a
S981002932	W	Ammonia by IC-Dionex	4000/4500	ug/g	107.1	<1.39e-01	<2.52e+03	<2.56e+03	n/a	n/a	93.17	2.52e+03	n/a
S981002932	W	Nitrate by IC-Dionex	4000/4500	ug/g	102.1	<1.20e-01	3.81e+03	6.56e+05	6.55e+05	0.15	109.8	2.81e+03	n/a
S981002932	W	Phosphate by IC-Dionex	4000/4500	ug/g	98.59	<1.38e-01	9.01e+03	<2.46e+03	9.23e+03	n/a	99.27	2.42e+03	n/a
S981002932	W	Sulfate by IC-Dionex	4000/4500	ug/g	108.3	<1.09e-01	3.14e+03	9.45e+03	3.23e+03	4.77	96.71	2.79e+03	n/a
S981002932	W	Oxalate by IC-Dionex	4000/4500	ug/g	108.3	<1.09e-01	3.14e+03	3.32e+03	3.23e+03	5.57	105.6	2.12e+03	n/a

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 250
SEGMENT #: 6

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002917		Bulk Density of Sample	g/mL	n/a	n/a	dry	n/a	n/a	n/a	n/a	5.00e-01	n/a
S98T002912		DSC Exotherm DRY Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S98T002912		DSC Exotherm on Perkin Elmer	Joules/g	99.02	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S98T002912		% Water by TGA on Perkin Elmer	%	99.17	n/a	1.070	1.550	1.310	36.6	n/a	n/a	n/a
S98T002912		TIC by Acid/Coulometry	ug/g	102.2	2.900	1.68e+03	1.77e+03	1.72e+03	5.22	86.50	5.000	n/a
S98T002912		TIC by Persulfate/Coulometry	ug/g	93.00	8.500	4.12e+02	4.510	4.315	9.04	80.50	40.00	n/a
S98T002920	F	Strontium-89/90 High Level	uCi/g	94.42	7.00e-03	5.95e-01	6.90e-01	6.43e-01	14.6	n/a	4.00e-03	3.67e+00
S98T002920	F	Aluminum - ICP-Fusion	ug/g	97.00	<5.00e-02	1.83e+03	1.81e+03	1.82e+03	1.10	94.80	1.00e+03	n/a
S98T002920	F	Boron - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.00e+03	<1.00e+03	n/a	n/a	98.70	1.00e+03	n/a
S98T002920	F	Barium - ICP-Fusion	ug/g	94.20	<5.00e-02	<2.01e+03	<2.01e+03	n/a	n/a	90.70	1.00e+03	n/a
S98T002920	F	Bismuth - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	97.40	2.01e+03	n/a
S98T002920	F	Calcium - ICP-Fusion	ug/g	105.0	<1.00e-01	<2.01e+03	<2.01e+03	n/a	n/a	106.0	2.01e+03	n/a
S98T002920	F	Chromium - ICP-Fusion	ug/g	101.0	<1.00e-02	<2.01e+02	225.0	n/a	n/a	100.0	201.0	n/a
S98T002920	F	Iron - ICP-Fusion	ug/g	96.60	<5.00e-02	<1.00e+03	<1.00e+03	n/a	n/a	95.10	1.00e+03	n/a
S98T002920	F	Lithium - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.01e+02	<2.01e+02	n/a	n/a	93.60	201.0	n/a
S98T002920	F	Manganese - ICP-Fusion	ug/g	98.80	<1.00e-02	<2.01e+02	<2.01e+02	n/a	n/a	95.70	201.0	n/a
S98T002920	F	Sodium - ICP-Fusion	ug/g	97.60	1.38e-01	2.43e+05	2.46e+05	2.45e+05	1.23	130.0	2.01e+03	n/a
S98T002920	F	Nickel - ICP-Fusion	ug/g	99.80	1.94e-01	3.93e+03	4.92e+03	4.42e+03	22.4	97.40	402.0	n/a
S98T002920	F	Silicon - ICP-Fusion	ug/g	104.0	<5.00e-02	<1.00e+03	<1.00e+03	n/a	n/a	106.0	1.00e+03	n/a
S98T002920	F	Uranium - ICP-Fusion	ug/g	94.70	<5.00e-01	<1.00e+04	<1.00e+04	n/a	n/a	90.00	1.00e+04	n/a
S98T002920	F	Zinc - ICP-Fusion	ug/g	96.20	<1.00e-02	<2.01e+02	<2.01e+02	n/a	n/a	96.00	201.0	n/a
S98T002920	F	Zirconium - ICP-Fusion	ug/g	96.20	<1.00e-02	<2.01e+02	<2.01e+02	n/a	n/a	94.00	201.0	n/a
S98T002920	F	Cesium-137 by GEA	uCi/g	99.76	<1.00e-02	7.543	7.890	7.716	4.54	n/a	n/a	1.02
S98T002920	F	Alpha of Digested Solid	uCi/g	91.11	<1.00e-01	<1.51e-01	<2.23e-1	n/a	n/a	86.91	3.04e-01	5.00e+02
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	109.0	<1.20e-03	<4.918	<4.358e-1	n/a	n/a	n/a	4.918	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	n/a	<1.20e-04	<4.92e-01	<4.35e-1	n/a	n/a	n/a	4.92e-01	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	n/a	<1.20e-04	<4.92e-01	<4.35e-1	n/a	n/a	n/a	4.92e-01	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	104.9	<1.20e-04	<4.92e-01	<4.35e-1	n/a	n/a	107.6	4.92e-01	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	n/a	<1.20e-04	<4.92e-01	<4.35e-1	n/a	n/a	n/a	6.35e-01	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	108.0	<1.20e-04	<4.92e-01	<4.35e-1	36.62	0.55	106.5	4.92e-01	n/a
S98T002928	A	Uranium-235 by ICP/MS Acid/GA505159	ug/g	92.19	1.66e-04	2.01e-01	2.14e-01	2.07e-01	6.27	n/a	8.20e-02	n/a
S98T002928	A	Neptunium-237 by ICP/MS	ug/g	118.4	<2.60e-05	<6.148	<5.735e-1	n/a	n/a	n/a	6.149	n/a
S98T002928	A	Neptunium-237 by ICP/MS	ug/g	100.6	<2.60e-05	<6.148e-01	<5.73e-1	n/a	n/a	n/a	9.70e-02	n/a
S98T002928	A	Plutonium-240 by ICP/MS	ug/g	n/a	<2.83e-05	<1.35e-01	<1.21e-1	n/a	n/a	n/a	1.35e-01	n/a
S98T002928	A	Plutonium-240 by ICP/MS	ug/g	n/a	<2.83e-05	<1.35e-01	<1.21e-1	n/a	n/a	n/a	1.35e-01	n/a
S98T002928	A	Plutonium-242 by ICP/MS	ug/g	104.3	<2.74e-05	<6.74e-02	<6.83e-2	n/a	n/a	n/a	6.70e-02	n/a
S98T002928	A	Pu/Am-241 by ICP/MS	ug/g	104.3	<2.74e-05	<1.02e-01	<1.53e-01	1.84e+03	n/a	n/a	1.02e-01	n/a
S98T002928	A	Aluminum - ICP-Acid Digest	ug/g	94.00	1.23e-01	1.42e+03	1.85e+03	1.84e+03	0.34	120.6	31.00	n/a
S98T002928	A	Boron - ICP-Acid Digest	ug/g	107.6	6.00e-01	1.42e+02	132.0	147.0	6.80	97.80	31.00	n/a
S98T002928	A	Barium - ICP-Acid Digest	ug/g	91.20	<1.00e-02	<1.510	<4.75e1	n/a	n/a	93.20	31.00	n/a
S98T002928	A	Bismuth - ICP-Acid Digest	ug/g	88.20	<1.00e-01	<1.02e+02	<9.50e1	n/a	n/a	86.60	102.0	n/a
S98T002928	A	Calcium - ICP-Acid Digest	ug/g	97.40	2.59e-01	1.42e+02	104.0	171.5	15.3	96.40	102.0	n/a
S98T002928	A	Chromium - ICP-Acid Digest	ug/g	92.80	<1.00e-02	1.74e+02	175.0	176.0	3.41	99.80	10.20	n/a
S98T002928	A	Iron - ICP-Acid Digest	ug/g	89.00	<1.00e-02	52.50	51.00	51.75	2.90	94.20	51.00	n/a

26-mar-1999 07:16:53

A-0002-1

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S981002928	A	Potassium - ICP-Acid Digest	ug/g	89.40	<5.00e-01	<5.10e-02	<4.75e2	N/A	N/A	141.8	510.0	N/A	N/A
S981002928	A	Lithium - ICP-Acid Digest	ug/g	94.40	<1.00e-02	<10.20	<9.50e0	N/A	N/A	95.40	10.20	N/A	N/A
S981002928	A	Manganese - ICP-Acid Digest	ug/g	90.60	<1.00e-02	<10.20	<9.50e0	N/A	N/A	95.60	10.20	N/A	N/A
S981002928	A	Sodium - ICP-Acid Digest	ug/g	116.6	9.87e-01	2.52e+05	2.55e+05	2.54e+05	1.18	1.56e+03	102.0	N/A	N/A
S981002928	A	Nickel - ICP-Acid Digest	ug/g	90.40	<2.00e-02	31.70	28.70	30.20	9.93	97.60	20.40	N/A	N/A
S981002928	A	Silicon - ICP-Acid Digest	ug/g	182.4	2.120	3.30e+02	214.0	272.0	42.6	147.2	51.00	N/A	N/A
S981002928	A	Uranium - ICP-Acid Digest	ug/g	90.60	<5.00e-01	<5.10e+02	<4.75e2	N/A	N/A	94.20	510.0	N/A	N/A
S981002928	A	Zinc - ICP-Acid Digest	ug/g	83.20	1.40e-02	<10.20	<9.50e0	N/A	N/A	96.20	10.20	N/A	N/A
S981002928	A	Zirconium - ICP-Acid Digest	ug/g	94.12	<6.20e-02	<75.71	<7.39e1	N/A	N/A	94.96	73.68	N/A	N/A
S981002933	M	Formate by IC-Dionex 4000/4500	ug/g	104.2	<1.20e-02	7.70e+02	723.0	746.4	6.30	105.9	121.3	N/A	N/A
S981002933	M	Fluoride by IC-Dionex 4000/4500	ug/g	98.87	<1.00e-01	3.99e+02	<3.41e2	N/A	N/A	93.72	340.2	N/A	N/A
S981002933	M	Chloride by IC-Dionex 4000/4500	ug/g	90.69	<1.00e-01	2.71e+03	3.01e+03	2.86e+03	10.5	87.59	2.16e+03	N/A	N/A
S981002933	M	Nitrite by IC-Dionex 4000/4500	ug/g	98.46	<1.25e-01	<2.50e+03	<2.51e3	N/A	N/A	93.00	2.50e+03	N/A	N/A
S981002933	M	Bromide by IC-Dionex 4000/4500	ug/g	101.5	<1.39e-01	7.18e+05	7.09e+05	7.14e+05	1.26	89.02	2.78e+03	N/A	N/A
S981002933	M	Nitrate by IC-Dionex 4000/4500	ug/g	96.90	<1.20e-01	<2.40e+03	<2.41e3	N/A	N/A	95.26	2.40e+03	N/A	N/A
S981002933	M	Phosphate by IC-Dionex 4000/4500	ug/g	98.28	<1.38e-01	3.61e+03	3.60e+03	3.61e+03	0.28	93.73	2.76e+03	N/A	N/A
S981002933	M	Sulfate by IC-Dionex 4000/4500	ug/g	109.2	<1.09e-01	<2.10e+03	<2.11e3	N/A	N/A	105.4	2.10e+03	N/A	N/A
S981002933	M	Oxalate by IC-Dionex 4000/4500	ug/g										

Table 1. Data Summary Report.
BY-105 (2)

PORTION: L Lower Half of Segment												
Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Error
S9810020208		Bulk Density of Sample	g/mL		N/A	1.890	N/A	N/A	N/A	N/A	5.00e-01	N/A
S9810020209		Bulk Density of Sample	Joules/g DRY		N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S9810020210		DSC Exotherm DRY Calculated	Joules/g DRY	99.02	N/A	0.00e+00	0.00e+00	0.00e+00	0.00	N/A	N/A	N/A
S9810020213		DSC Exotherm on Perkin Elmer	Joules/g	99.17	N/A	10.28	12.31	11.29	18.0	N/A	N/A	N/A
S9810020213		% Water by TGA on Perkin Elmer	%	102.2	2.900	1.30e+03	1.32e+03	1.31e+03	1.53	N/A	5.000	N/A
S9810020215		TIC by Acid/Countometry	uV/g	93.40	8.500	6.65e+02	631.0	648.0	5.25	N/A	40.00	N/A
S9810020213		TIC by Percent (rate/Countometry	uV/g	94.42	7.00e-03	6.43e-01	7.47e-01	6.99e-01	15.0	N/A	4.00e-03	3.50E+00
S9810020221		F Strontium-89/90 High Level	uV/g	97.00	5.00e-02	7.13e+03	7.39e+03	7.26e+03	3.58	N/A	1.00e+03	N/A
S9810020221		F Aluminum - ICP-Fusion	uV/g	101.6	5.00e-02	<1.01e+03	<1.01e3	N/A	N/A	N/A	1.00e+03	N/A
S9810020221		F Boron - ICP-Fusion	uV/g	94.20	5.00e-02	<1.01e+03	<1.01e3	N/A	N/A	N/A	1.00e+03	N/A
S9810020221		F Barium - ICP-Fusion	uV/g	100.0	<1.00e-01	<2.01e+03	<2.03e3	N/A	N/A	N/A	2.01e+03	N/A
S9810020221		F Bismuth - ICP-Fusion	uV/g	101.0	<1.00e-01	<2.01e+03	<2.03e3	N/A	N/A	N/A	2.01e+03	N/A
S9810020221		F Chromium - ICP-Fusion	uV/g	101.0	<1.00e-02	3.75e+02	384.0	379.5	2.37	N/A	201.0	N/A
S9810020221		F Iron - ICP-Fusion	uV/g	98.60	5.00e-02	<1.01e+03	<1.01e3	N/A	N/A	N/A	1.00e+03	N/A
S9810020221		F Lithium - ICP-Fusion	uV/g	94.20	<1.00e-02	<2.01e+03	<2.03e2	N/A	N/A	N/A	201.0	N/A
S9810020221		F Lithium - ICP-Fusion	uV/g	97.80	<1.00e-02	<2.01e+03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S9810020221		F Manganese - ICP-Fusion	uV/g	97.60	1.38e-01	2.30e+05	2.30e+05	2.30e+05	0.00	N/A	2.01e+03	N/A
S9810020221		F Sodium - ICP-Fusion	uV/g	99.80	1.38e-01	7.29e+03	8.00e+03	7.64e+03	9.29	N/A	402.0	N/A
S9810020221		F Nickel - ICP-Fusion	uV/g	104.0	5.00e-02	<1.01e+03	<1.01e3	N/A	N/A	N/A	1.00e+03	N/A
S9810020221		F Silicon - ICP-Fusion	uV/g	94.70	5.00e-01	<1.01e+04	<1.01e4	N/A	N/A	N/A	1.00e+03	N/A
S9810020221		F Uranium - ICP-Fusion	uV/g	96.20	<1.00e-02	<2.01e+03	<2.03e2	N/A	N/A	N/A	201.0	N/A
S9810020221		F Zinc - ICP-Fusion	uV/g	96.20	<1.00e-02	<2.01e+03	<2.03e2	N/A	N/A	N/A	201.0	N/A
S9810020221		F Zirconium - ICP-Fusion	uV/g	99.76	<1.80e-02	33.32	35.70	33.51	1.19	N/A	N/A	0.480
S9810020221		F Cesium-137 by GEA	uV/g	91.11	<1.14e-01	<1.88e-01	<1.15E-1	N/A	N/A	N/A	3.00e+02	N/A
S9810020221		F Alpha of Digested Solid	uV/g	109.0	<1.20e-03	4.744	44.90E0	N/A	N/A	N/A	4.744	N/A
S9810020229		A 1129 by ICP/MS Actinide/LA505159	uV/g	N/A	<1.20e-04	44.74e-01	44.90e-1	N/A	N/A	N/A	4.74e-01	N/A
S9810020229		A Uranium-233 by ICP/MS Actinide/LA505159	uV/g	N/A	<1.20e-04	44.74e-01	44.90e-1	N/A	N/A	N/A	4.74e-01	N/A
S9810020229		A Uranium-234 by ICP/MS Actinide/LA505159	uV/g	104.9	<1.20e-04	44.74e-01	44.90e-1	N/A	N/A	N/A	109.7	4.74e-01
S9810020229		A Uranium-235 by ICP/MS Actinide/LA505159	uV/g	N/A	<1.60e-04	46.33e-01	46.54e-1	N/A	N/A	N/A	6.32e-01	N/A
S9810020229		A Uranium-238 by ICP/MS Actinide/LA505159	uV/g	108.0	<1.20e-04	22.11	23.60	22.86	6.56	110.5	7.90e-02	N/A
S9810020229		A Uranium-238 by ICP/MS Actinide/LA505159	uV/g	92.19	1.76e-04	9.33e-01	9.44e-1	9.39e-1	1.17	N/A	7.90e-02	N/A
S9810020229		A Tc99 by ICP/MS Actinide/LA505159	uV/g	N/A	<1.65e-03	<5.930	<6.13e0	N/A	N/A	N/A	5.930	N/A
S9810020229		A Sn126 ICP/MS Actinide/LA-505159	uV/g	118.4	<2.60e-05	<9.35e-02	<9.66e-2	N/A	N/A	N/A	9.40e-02	N/A
S9810020229		A Neptunium-237 by ICP/MS	uV/g	N/A	<3.64e-05	<1.31e-01	<1.35e-1	N/A	N/A	N/A	1.31e-01	N/A
S9810020229		A Plutonium-239 by ICP/MS	uV/g	100.6	<2.83e-05	<1.02e-01	<1.09e-1	N/A	N/A	N/A	1.02e-01	N/A
S9810020229		A Plutonium-240 by ICP/MS	uV/g	N/A	<1.81e-05	<6.50e-02	<6.72e-2	N/A	N/A	N/A	6.50e-02	N/A
S9810020229		A Plutonium-242 by ICP/MS	uV/g	104.3	<2.74e-05	<9.87e-02	<1.02e-1	N/A	N/A	N/A	9.90e-02	N/A
S9810020229		A Pu/Am-241 by ICP/MS	uV/g	94.0	1.23e-01	8.27e+03	8.39e+03	8.33e+03	1.44	N/A	49.10	N/A
S9810020229		A Aluminum - ICP-Acid Digest	uV/g	107.6	6.08e-01	1.25e+02	156.0	140.5	22.1	N/A	49.10	N/A
S9810020229		A Boron - ICP-Acid Digest	uV/g	91.80	<5.00e-02	<49.20	<5.08e1	N/A	N/A	N/A	49.10	N/A
S9810020229		A Barium - ICP-Acid Digest	uV/g	98.80	<1.00e-01	<98.30	<1.02e2	N/A	N/A	N/A	98.30	N/A
S9810020229		A Bismuth - ICP-Acid Digest	uV/g	97.40	<2.59e-01	2.11e+02	107.0	159.0	65.4	N/A	98.50	N/A
S9810020229		A Calcium - ICP-Acid Digest	uV/g	92.80	<1.00e-02	3.69e+02	382.0	375.5	3.46	N/A	9.800	N/A
S9810020229		A Chromium - ICP-Acid Digest	uV/g	89.00	<5.00e-02	86.70	58.60	72.65	38.7	N/A	49.10	N/A
S9810020229		A Iron - ICP-Acid Digest	uV/g									

26-mar-1999 07:17:58

A-0002-1

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002929	A	Potassium -ICP-Acid Digest	ug/g	89.40	<5.00e-01	1.04e+03	766.0	903.0	30.3	N/A	9.92E-01	N/A
S981002929	A	Lithium -ICP-Acid Digest	ug/g	94.40	<1.00e-02	< 9.830	<1.02e1	N/A	N/A	N/A	9.850	N/A
S981002929	A	Manganese -ICP-Acid Digest	ug/g	90.60	<1.00e-02	< 9.830	<1.02e1	N/A	N/A	N/A	9.850	N/A
S981002929	A	Sodium -ICP-Acid Digest	ug/g	116.6	9.87e-01	2.43e+05	2.42e+05	2.43e+05	0.41	N/A	98.50	N/A
S981002929	A	Nickel -ICP-Acid Digest	ug/g	90.40	<2.00e-02	20.30	26.20	23.25	25.4	N/A	19.70	N/A
S981002929	A	Silicon -ICP-Acid Digest	ug/g	182.4	2.120	2.34e+02	226.0	230.0	3.48	N/A	49.10	N/A
S981002929	A	Uranium -ICP-Acid Digest	ug/g	83.20	<5.00e-01	<4.92e+02	<5.08e2	N/A	N/A	N/A	9.850	N/A
S981002929	A	Zinc -ICP-Acid Digest	ug/g	89.60	<1.00e-02	< 9.830	<1.02e1	N/A	N/A	N/A	9.850	N/A
S981002929	A	Chromium (VI) by Spec.	ug/g	100.8	<27.51	2.25e+02	208.0	216.3	7.83	N/A	1.00e-02	N/A
S981002934	U	Formate by IC-Dionex 4000/4500	ug/g	94.12	<6.20e-02	2.06e+02	183.0	194.3	1.83	N/A	72.66	N/A
S981002934	U	Fluoride-IC-Dionex 4000/4500	ug/g	104.2	<1.20e-02	8.35e+02	698.0	766.3	11.9	106.1	119.6	N/A
S981002934	U	Chloride-IC-Dionex 4000/4500	ug/g	98.87	<1.70e-02	9.12e+02	902.0	907.2	1.10	N/A	335.5	N/A
S981002934	U	Nitrite-IC - Dionex 4000/4500	ug/g	90.69	<1.08e-01	1.20e+04	1.07e+04	1.14e+04	11.5	N/A	2.13e+03	N/A
S981002934	U	Bromide by Ion Chromatograph	ug/g	98.46	<1.25e-01	<2.47e+03	<2.40e3	N/A	N/A	N/A	2.74e+03	N/A
S981002934	U	Nitrate by IC-Dionex 4000/4500	ug/g	101.5	<1.20e-01	6.26e+05	6.20e+05	6.23e+05	0.96	N/A	2.74e+03	N/A
S981002934	U	Phosphate-IC-Dionex 4000/4500	ug/g	96.90	<1.38e-01	<2.37e+03	<2.30e3	3.15e+03	5.08	N/A	2.72e+03	N/A
S981002934	U	Sulfate by IC-Dionex 4000/4500	ug/g	98.28	<1.38e-01	<2.37e+03	3.15e+03	3.15e+03	5.08	N/A	2.72e+03	N/A
S981002934	U	Oxalate-IC-Dionex 4000/4500	ug/g	109.2	<1.05e-01	<2.07e+03	<2.09e3	N/A	N/A	N/A	2.07e+03	N/A

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 250
SEGMENT #: 98

SEGMENT PORTION: L Lower Half of Segment

Sample#	RAM	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002909		Bulk Density of Sample	g/ml	n/a	n/a	DRY	n/a	n/a	n/a	n/a	5.00e-01	n/a
S981002914		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002914		DSC Exotherm on Perkin Elmer	Joules/g	99.02	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002914		% Water by TOA on Perkin Elmer	%	99.17	n/a	15.86	14.90	15.38	6.24	n/a	n/a	n/a
S981002914		TIC by Acid/Coulometry	ug/g	99.50	4.500	1.36e+04	2.27e+04	2.27e+04	80.2	69.10	5.00	n/a
S981002914		TIC by Persulfate/Coulometry	ug/g	94.00	2.100	7.84e+03	6.50e+03	7.17e+03	18.7	119.0	40.00	n/a
S981002922		TOC by Persulfate/High Level	ug/g	94.42	7.00e-03	3.42e+02	159.0	250.5	73.1	n/a	9.70e-02	7.72E-01
S981002922		Aluminum - ICP-Fusion	ug/g	97.00	<5.00e-02	1.71e+04	1.49e+04	1.60e+04	13.8	90.30	1.05e+03	n/a
S981002922		Boron - ICP-Fusion	ug/g	101.6	<5.00e-02	<1.05e+03	<1.04e3	n/a	n/a	94.50	1.05e+03	n/a
S981002922		Beryllium - ICP-Fusion	ug/g	94.20	<1.00e-01	<2.95e+03	<2.08e3	n/a	n/a	86.60	1.05e+03	n/a
S981002922		Bismuth - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.95e+03	<2.08e3	n/a	n/a	98.10	2.09e+03	n/a
S981002922		Calcium - ICP-Fusion	ug/g	105.0	<1.00e-01	<2.95e+03	<2.08e3	2.82e+04	9.59	105.0	2.09e+03	n/a
S981002922		Chromium - ICP-Fusion	ug/g	101.0	<1.00e-02	7.68e+02	786.0	777.0	2.32	97.60	209.0	n/a
S981002922		Cobalt - ICP-Fusion	ug/g	96.60	<5.00e-02	1.29e+04	7.41e+03	1.02e+04	54.1	92.70	1.05e+03	n/a
S981002922		Lithium - ICP-Fusion	ug/g	98.20	<1.00e-02	<2.09e+02	<2.08e2	n/a	n/a	86.10	209.0	n/a
S981002922		Manganese - ICP-Fusion	ug/g	98.80	<1.00e-02	2.10e+02	<2.08e2	1.86e+05	8.09	78.00	2.09e+03	n/a
S981002922		Selenium - ICP-Fusion	ug/g	97.60	1.38e-01	1.78e+04	1.25e+04	1.22e+04	5.76	97.70	418.0	n/a
S981002922		Silicon - ICP-Fusion	ug/g	99.80	<5.00e-02	5.33e+03	6.08e+03	5.70e+03	13.1	103.0	1.05e+03	n/a
S981002922		Titanium - ICP-Fusion	ug/g	94.70	<5.00e-01	2.26e+04	1.14e+04	1.70e+04	65.9	86.00	1.05e+04	n/a
S981002922		Zinc - ICP-Fusion	ug/g	96.20	<1.00e-02	<2.09e+02	<2.08e2	n/a	n/a	96.70	209.0	n/a
S981002922		Zirconium - ICP-Fusion	ug/g	96.20	<1.00e-02	<2.09e+02	<2.08e2	n/a	n/a	90.50	209.0	n/a
S981002922		Zirconium - ICP-Fusion	ug/g	99.76	<1.80e-02	70.38	61.50	65.94	13.5	n/a	n/a	1.08
S981002922		Aluminum - ICP-Fusion	ug/g	91.11	<1.14e-01	<2.93e+03	2.04e+01	n/a	n/a	84.69	3.16e-01	1.59E+02
S981002930		Aluminum - ICP-Fusion	ug/g	109.0	<1.20e-03	<4.74e+02	<4.81e2	n/a	n/a	n/a	474.6	n/a
S981002930		Uranium - ICP-Fusion	ug/g	n/a	<1.20e-04	<4.68e9	<4.76e9	n/a	n/a	n/a	4.689	n/a
S981002930		Uranium - ICP-Fusion	ug/g	n/a	<1.20e-04	<4.68e9	<4.76e9	n/a	n/a	n/a	4.689	n/a
S981002930		Uranium - ICP-Fusion	ug/g	104.2	<1.20e-04	<4.68e9	<4.76e9	n/a	n/a	96.34	4.689	n/a
S981002930		Uranium - ICP-Fusion	ug/g	104.0	<1.60e-04	<6.35e9	<6.35e9	n/a	n/a	104.5	6.251	n/a
S981002930		Uranium - ICP-Fusion	ug/g	104.0	<1.20e-04	1.84e+04	1.18e+04	1.51e+04	43.7	104.5	6.689	n/a
S981002930		Uranium - ICP-Fusion	ug/g	92.19	<1.76e-04	<3.204	<3.26e9	n/a	n/a	n/a	3.204	n/a
S981002930		Uranium - ICP-Fusion	ug/g	118.4	<1.65e-03	<2.40e+02	<2.44e2	n/a	n/a	112.3	3.790	n/a
S981002930		Uranium - ICP-Fusion	ug/g	100.6	<2.64e-05	<5.308	<5.39e9	n/a	n/a	96.12	5.308	n/a
S981002930		Uranium - ICP-Fusion	ug/g	n/a	<2.81e-05	<4.124	<4.19e9	n/a	n/a	n/a	4.124	n/a
S981002930		Uranium - ICP-Fusion	ug/g	104.3	<1.85e-05	<2.636	<2.68e9	n/a	n/a	n/a	2.635	n/a
S981002930		Uranium - ICP-Fusion	ug/g	104.3	<2.74e-05	<3.997	<4.06e9	n/a	n/a	100.9	3.998	n/a
S981002930		Uranium - ICP-Fusion	ug/g	94.00	1.23e-01	1.50e+04	1.41e+04	1.46e+04	6.19	84.20	48.60	n/a
S981002930		Uranium - ICP-Fusion	ug/g	107.6	<6.08e-01	1.42e+02	115.0	128.5	21.0	96.80	48.60	n/a
S981002930		Uranium - ICP-Fusion	ug/g	91.80	<5.00e-02	1.18e+02	98.10	108.0	18.4	92.60	48.60	n/a
S981002930		Uranium - ICP-Fusion	ug/g	88.80	<1.00e-01	9.42e+02	643.0	792.5	37.7	67.00	97.00	n/a
S981002930		Uranium - ICP-Fusion	ug/g	97.40	<2.59e-01	2.65e+04	2.30e+04	2.48e+04	14.1	7.62e02	97.00	n/a
S981002930		Uranium - ICP-Fusion	ug/g	92.80	<1.00e-02	6.15e+02	624.0	619.5	1.45	104.2	9.700	n/a
S981002930		Uranium - ICP-Fusion	ug/g	89.00	<5.00e-02	1.02e+04	7.09e+03	8.62e+03	36.5	-2.82e02	48.60	n/a

Sample#	R	MF	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002930	A	Potassium	-ICP-Acid Digest	ug/g	89.40	<5.00e-01	9.59e+02	1.08e+03	1.07e+03	20.7	98.00	486.0	n/a
S98T002930	A	Lithium	-ICP-Acid Digest	ug/g	94.40	<1.00e-02	9.720	9.780	122.5	36.7	91.20	9.700	n/a
S98T002930	A	Manganese	-ICP-Acid Digest	ug/g	90.60	<1.00e-02	4.55e+02	2.27e+03	2.23e+03	0.90	1.950e+03	97.00	n/a
S98T002930	A	Sodium	-ICP-Acid Digest	ug/g	116.6	9.87e-01	2.26e+03	3.27e+03	3.20e+03	47.4	1.370e+02	19.40	n/a
S98T002930	A	Nickel	-ICP-Acid Digest	ug/g	10.40	<2.00e-02	3.79e+03	4.25e+03	3.76e+03	26.1	212.0	48.60	n/a
S98T002930	A	Silicon	-ICP-Acid Digest	ug/g	82.4	<5.00e-02	1.57e+04	1.01e+04	1.29e+04	43.4	2.120e+02	486.0	n/a
S98T002930	A	Uranium	-ICP-Acid Digest	ug/g	80.60	<5.00e-02	4.85e+03	88.30	68.0	58.2	95.60	9.700	n/a
S98T002930	A	Zinc	-ICP-Acid Digest	ug/g	87.60	<1.00e-02	9.720	9.780	122.5	36.7	91.20	9.700	n/a
S98T002935	W	Chromium (VI)	by Spec.	ug/g	87.60	<1.00e-01	1.75e+02	203.0	189.5	14.2	96.60	1.00e-02	n/a
S98T002935	W	Formate by IC-Dioxin	4000/4500	ug/g	100.5	<5.00e-02	2.45e+02	303.0	273.8	21.2	100.5	75.00	n/a
S98T002935	W	Fluoride by IC-Dioxin	4000/4500	ug/g	104.5	<1.00e-02	7.49e+02	1.04e+03	893.1	32.9	102.5	123.4	n/a
S98T002935	W	Chloride by IC-Dioxin	4000/4500	ug/g	99.12	<1.00e-02	8.75e+02	1.39e+03	1.10e+03	41.3	96.36	176.8	n/a
S98T002935	W	Nitrite by Ion Chromatograph	4500	ug/g	89.15	<1.00e-01	1.12e+04	1.39e+04	1.26e+04	21.5	88.43	1.11e+03	n/a
S98T002935	W	Bromide by Ion Chromatograph	4500	ug/g	99.15	<1.00e-01	1.28e+03	3.04e+05	3.56e+05	29.0	149.0	1.28e+03	n/a
S98T002935	W	Nitrate by IC-Dioxin	4000/4500	ug/g	108.3	<1.39e-01	4.70e+03	2.82e+03	2.82e+03	30.7	97.08	1.23e+03	n/a
S98T002935	W	Phosphate by IC-Dioxin	4000/4500	ug/g	97.81	<1.38e-01	6.73e+03	9.20e+03	7.98e+03	30.7	101.1	1.42e+03	n/a
S98T002935	W	Sulfate by IC-Dioxin	4000/4500	ug/g	101.1	<1.38e-01	6.73e+03	9.20e+03	7.98e+03	30.7	101.1	1.42e+03	n/a
S98T002935	W	Oxalate by IC-Dioxin	4000/4500	ug/g	103.4	<1.05e-01	7.86e+03	1.20e+04	9.93e+03	41.7	103.0	1.08e+03	n/a

Drainable Liquid: Drainable Liquid

Sample#	R	MF	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002938	A	Exotherm	Dry Calculated	Joules/g Dry	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S98T002938	A	Exotherm	on Perkin-Elmer	Joules/g	98.91	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S98T002938	A	Exotherm	by ISE-Rid Additions	ug/mL	95.06	<5.000	55.90	61.40	58.65	9.38	93.90	5.000	n/a
S98T002938			Specific Gravity	Sp. G.	98.92	n/a	1.467	1.457	1.462	0.68	n/a	1.00e-03	n/a
S98T002938			% Water by IGA on Perkin-Elmer	%	99.17	n/a	52.02	48.06	50.04	7.91	n/a	n/a	n/a
S98T002938			Alpha in Liquid Samples	uCi/mL	97.78	<1.26e-02	1.49e-02	1.26E-2	1.63E-2	n/a	94.81	1.90e-02	2.11E+02
S98T002938			Strontium-89/90 High Level	uCi/mL	101.9	<2.00e-03	1.62e-02	1.63E-2	1.63E-2	0.62	n/a	2.00e-03	1.65E+01
S98T002938			Pu-239/240 by TRU-SPEC Resin	uCi/mL	105.4	<5.35e-04	3.12e-04	4.2E-4	4.88E-4	1.64	n/a	3.12e-04	9.25E+00
S98T002939	D	Aluminum	-ICP-Acid Dil.	ug/mL	97.80	<5.00e-02	4.92e+04	4.84e+04	4.88e+04	1.64	2.860e+01	30.10	n/a
S98T002939	D	Boron	-ICP-Acid Dil.	ug/mL	102.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	102.0	30.10	n/a
S98T002939	D	Barium	-ICP-Acid Dil.	ug/mL	96.40	<5.00e-02	< 30.10	<6.01e1	n/a	n/a	90.60	30.10	n/a
S98T002939	D	Bismuth	-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	104.0	60.10	n/a
S98T002939	D	Calcium	-ICP-Acid Dil.	ug/mL	103.2	<1.00e-01	< 60.10	<6.01e1	n/a	n/a	99.70	60.10	n/a
S98T002939	D	Chromium	-ICP-Acid Dil.	ug/mL	100.0	<1.00e-02	1.03e+03	1.03e+03	1.03e+03	0.00	96.60	6.010	n/a
S98T002939	D	Iron	-ICP-Acid Dil.	ug/mL	96.40	<5.00e-02	1.25e+02	124.0	124.5	0.80	92.30	30.10	n/a
S98T002939	D	Lithium	-ICP-Acid Dil.	ug/mL	97.20	<1.00e-02	4.81e+03	4.57e+03	4.69e+03	5.12	94.50	300.0	n/a
S98T002939	D	Potassium	-ICP-Acid Dil.	ug/mL	100.4	<1.00e-02	6.010	<6.01e0	n/a	n/a	91.70	60.10	n/a
S98T002939	D	Lithium	-ICP-Acid Dil.	ug/mL	98.40	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	93.60	6.010	n/a
S98T002939	D	Sodium	-ICP-Acid Dil.	ug/mL	99.80	<1.00e-01	2.15e+05	2.08e+05	2.12e+05	3.31	230e2	60.10	n/a
S98T002939	D	Nickel	-ICP-Acid Dil.	ug/mL	106.20	<2.00e-02	< 12.00	<203.1	221.5	16.7	98.50	12.00	n/a
S98T002939	D	Silicon	-ICP-Acid Dil.	ug/mL	98.40	<5.00e-02	2.40e+02	<3.01e2	n/a	n/a	91.50	300.0	n/a
S98T002939	D	Uranium	-ICP-Acid Dil.	ug/mL	96.10	<5.00e-02	<3.01e2	<3.01e2	n/a	n/a	93.70	6.010	n/a
S98T002939	D	Zinc	-ICP-Acid Dil.	ug/mL	96.60	<1.00e-02	< 6.010	<6.01e0	n/a	n/a	93.10	6.010	n/a
S98T002939	D	Zirconium	-ICP-Acid Dil.	ug/mL	107.6	<2.00e-02	< 6.010	<6.01e0	n/a	n/a	114.3	37.57	n/a
S98T002939	D	Formate by Ion Chromatograph	4500	ug/mL	99.69	<1.52e-03	2.33e+02	227.0	230.0	2.61	n/a	n/a	0.110
S98T002939	D	Chromium	137 by GEA	uCi/mL	102.4	<2.58e-01	9.41e+02	931.0	936.0	1.07	113.0	1.00e-02	n/a
S98T002940	D	Chromium (VI)	by Spec.	ug/mL	105.0	<1.20e-04	< 4.860	<4.860	n/a	n/a	n/a	4.860	n/a
S98T002940	D	I 129	by ICP/MS Acid Addition	ug/mL	105.0	<1.20e-04	< 4.860	<4.860	n/a	n/a	n/a	4.860	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	Spk. Rec. %	Det Limit	Count Err
S98T002940		OH- by Pot. Titration	ug/mL	100.0	n/a	5.08e-06	5.08e-06	5.08e-06	6.84	2.50e-03	n/a
S98T002940		TOC by Persulfate/Coulometry	ug/mL	92.00	9.600	2.72e-03	2.68e-03	2.70e-03	1.48	101.0	4.00
S98T002940	D	Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-03	<4.86e-01	<4.86e-01	n/a	n/a	4.86e-01	n/a
S98T002940	D	Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-03	<4.86e-01	<4.86e-01	n/a	n/a	4.86e-01	n/a
S98T002940	D	Uranium-235 by ICP/MS Acid Add	ug/mL	100.6	<1.20e-03	<4.86e-01	<4.86e-01	n/a	100.6	4.86e-01	n/a
S98T002940	D	Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<1.60e-03	<6.48e-01	<6.48e-01	n/a	n/a	6.48e-01	n/a
S98T002940	D	Uranium-238 by ICP/MS Acid Add	ug/mL	101.0	<1.20e-03	2.531	2.510	2.571	3.11	98.50	4.86e-01
S98T002940	D	Tc 99 by ICP/MS Acid Addition	ug/mL	78.48	<2.00e-06	4.212	4.160	4.186	1.19	81.28	8.10e-02
S98T002940	D	Sn 126 by ICP/MS Acid Addition	ug/mL	n/a	<1.50e-04	<6.075	<6.080	n/a	n/a	6.075	n/a
S98T002940	A	Neptunium-237 by ICP/MS	ug/mL	134.1	n/a	<6.53e-01	<6.53e-01	n/a	154.2	6.53e-01	n/a
S98T002940	A	Plutonium-239 by ICP/MS	ug/mL	139.0	<1.75e-03	<7.09e-01	<7.09e-01	n/a	154.2	7.09e-01	n/a
S98T002940	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<1.02e-03	<4.14e-01	<4.14e-01	n/a	n/a	4.14e-01	n/a
S98T002940	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<2.39e-03	<9.68e-01	<9.68e-01	n/a	n/a	9.68e-01	n/a
S98T002940	A	Pu/Am-241 by ICP/MS	ug/mL	n/a	<1.58e-03	<6.41e-01	<6.41e-01	n/a	n/a	6.41e-01	n/a

Table 1. Data Summary Report.
BY-105 (2)

CORE NUMBER: 250
SEGMENT #: Composite

SEGMENT PORTION: Core Composite

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S931003580		Bulk Density of Sample	g/mL	n/a	n/a	DRY	n/a	n/a	n/a	n/a	5.00e-01	n/a

Table 1. Data Summary Report.
BY-105 (2)

CORE NUMBER: 250

SEGMENT #: Core Composite

SEGMENT PORTION: Core Composite

Sample#	RAP	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003386		DSC Exotherm Dry, Calculated	loutles/g Dry	n/a	n/a	45.31	39.40	42.36	14.0	n/a	n/a	n/a
S98T003386		DSC Exotherm on Perkin Elmer	loutles/g	96.59	n/a	43.74	38.04	40.89	13.9	n/a	n/a	n/a
S98T003386		% Water by TGA on Perkin Elmer	%	102.1	n/a	3.420	3.500	3.460	2.31	n/a	n/a	n/a
S98T003386		TIC by Acid/Coulometry	ug/g	90.36	4.500	3.19e+03	3.40e+03	3.30e+03	6.37	105.0	5.000	n/a
S98T003386		TIC by Persulfate/Coulometry	ug/g	93.33	15.10	1.18e+03	1.26e+03	1.22e+03	6.56	95.20	40.00	n/a
S98T003386		TOC by Persulfate/Coulometry	ug/g	97.40	3.80e-02	1.420	1.320	1.370	7.30	n/a	2.00e-03	1.21e+00
S98T003388		F Strontium-89/90 High Level	uc/g	99.04	1.50e+04	1.61e+04	1.50e+04	1.56e+04	7.07	93.30	1.01e+03	n/a
S98T003388		F Aluminum - ICP-Fusion	ug/g	97.40	5.00e-02	1.61e+03	1.01e+03	n/a	n/a	103.0	1.01e+03	n/a
S98T003388		F Boron - ICP-Fusion	ug/g	107.8	5.00e-02	1.61e+03	1.01e+03	n/a	n/a	90.00	1.01e+03	n/a
S98T003388		F Barium - ICP-Fusion	ug/g	96.40	5.00e-02	1.61e+03	1.01e+03	n/a	n/a	96.30	2.03e+03	n/a
S98T003388		F Bismuth - ICP-Fusion	ug/g	100.4	1.00e-01	2.03e+03	2.03e+03	n/a	n/a	101.0	2.03e+03	n/a
S98T003388		F Calcium - ICP-Fusion	ug/g	101.8	1.00e-01	2.03e+03	2.03e+03	n/a	n/a	98.50	203.0	n/a
S98T003388		F Chromium - ICP-Fusion	ug/g	99.60	1.00e-02	3.36e+02	317.0	326.5	5.82	94.50	1.01e+03	n/a
S98T003388		F Iron - ICP-Fusion	ug/g	95.80	5.00e-02	1.61e+03	1.01e+03	n/a	n/a	94.50	1.01e+03	n/a
S98T003388		F Lithium - ICP-Fusion	ug/g	99.80	1.00e-02	2.03e+02	2.03e+02	n/a	n/a	90.20	203.0	n/a
S98T003388		F Manganese - ICP-Fusion	ug/g	98.00	1.00e-02	2.03e+02	2.03e+02	n/a	n/a	95.40	203.0	n/a
S98T003388		F Sodium - ICP-Fusion	ug/g	99.20	1.00e-01	2.42e+05	2.33e+05	2.38e+05	3.79	58.60	2.03e+03	n/a
S98T003388		F Nickel - ICP-Fusion	ug/g	98.80	7.40e-02	8.52e+02	1.47e+03	1.16e+03	55.2	97.80	405.0	n/a
S98T003388		F Silicon - ICP-Fusion	ug/g	107.4	5.00e-02	1.03e+03	1.07e+03	1.05e+03	3.81	108.0	1.01e+03	n/a
S98T003388		F Uranium - ICP-Fusion	ug/g	96.20	1.00e-02	2.03e+02	2.03e+02	n/a	n/a	91.50	1.01e+04	n/a
S98T003388		F Zinc - ICP-Fusion	ug/g	101.4	1.00e-02	2.03e+02	2.03e+02	n/a	n/a	97.50	203.0	n/a
S98T003388		F Zirconium - ICP-Fusion	ug/g	97.44	1.09e-02	16.52	15.80	16.16	4.33	n/a	n/a	0.690
S98T003388		F Alpha of Digested Solid	uc/g	93.87	1.50e-01	1.15e-01	2.26e-01	n/a	n/a	85.19	2.00e-01	1.40e+02
S98T003389		F Alpha of Digested Solid	uc/g	109.0	1.20e-03	4.945	4.58e0	n/a	n/a	n/a	4.945	n/a
S98T003389		F Uranium-233 by ICP/MS	ug/g	n/a	1.20e-04	4.94e-01	4.58e-1	n/a	n/a	n/a	4.94e-01	n/a
S98T003389		F Uranium-234 by ICP/MS	ug/g	102.8	1.20e-04	4.94e-01	4.58e-1	6.23e-01	4.33	103.4	4.94e-01	n/a
S98T003389		F Uranium-235 by ICP/MS	ug/g	103.0	1.20e-04	4.94e-01	4.58e-1	6.23e-01	4.33	103.4	4.94e-01	n/a
S98T003389		F Uranium-238 by ICP/MS	ug/g	103.0	1.20e-04	4.94e-01	4.58e-1	6.23e-01	4.33	103.4	4.94e-01	n/a
S98T003389		F Uranium-238 by ICP/MS	ug/g	93.66	1.20e-04	4.94e-01	4.58e-1	6.23e-01	4.33	103.4	4.94e-01	n/a
S98T003389		F Tc99 by ICP/MS	ug/g	93.66	1.20e-04	4.94e-01	4.58e-1	6.23e-01	4.33	103.4	4.94e-01	n/a
S98T003389		F Sn126 by ICP/MS	ug/g	119.2	4.72e-05	6.181	5.72e0	4.56e-01	4.39	95.43	8.20e-02	n/a
S98T003389		F Neptunium-237 by ICP/MS	ug/g	119.2	4.72e-05	6.181	5.72e0	4.56e-01	4.39	95.43	8.20e-02	n/a
S98T003389		F Plutonium-239 by ICP/MS	ug/g	102.4	3.47e-05	1.92e-01	1.78e-1	n/a	n/a	115.4	1.92e-01	n/a
S98T003389		F Plutonium-240 by ICP/MS	ug/g	n/a	3.47e-05	1.92e-01	1.78e-1	n/a	n/a	99.60	1.43e-01	n/a
S98T003389		F Plutonium-242 by ICP/MS	ug/g	n/a	3.47e-05	1.92e-01	1.78e-1	n/a	n/a	99.60	1.43e-01	n/a
S98T003389		F Pu/Am-241 by ICP/MS	ug/g	107.8	7.72e-05	1.53e-01	1.41e-1	n/a	n/a	104.3	1.53e-01	n/a
S98T003389		F Aluminum - ICP-Acid Digest	ug/g	94.80	1.20e-01	1.56e+04	1.50e+04	1.53e+04	3.92	226.0	51.50	n/a
S98T003389		F Boron - ICP-Acid Digest	ug/g	106.0	7.41e-01	1.06e+02	1.38e+02	124.5	21.7	104.6	51.50	n/a
S98T003389		F Barium - ICP-Acid Digest	ug/g	93.60	5.00e-02	51.30	49.48e1	n/a	n/a	92.40	51.50	n/a
S98T003389		F Bismuth - ICP-Acid Digest	ug/g	97.80	1.65e-01	1.21e+02	98.20	109.6	20.8	104.8	102.0	n/a
S98T003389		F Calcium - ICP-Acid Digest	ug/g	91.40	1.00e-02	2.83e+02	277.0	280.0	2.14	101.6	10.20	n/a
S98T003389		F Chromium - ICP-Acid Digest	ug/g	94.00	4.00e-02	1.69e+02	151.0	160.0	11.2	95.20	51.50	n/a
S98T003389		F Iron - ICP-Acid Digest	ug/g	90.60	5.00e-01	5.13e+02	44.74e2	n/a	n/a	146.8	515.0	n/a
S98T003389		F Potassium - ICP-Acid Digest	ug/g	90.60	5.00e-01	5.13e+02	44.74e2	n/a	n/a	146.8	515.0	n/a

26-mar-1999 07:12:42
A-0002-1

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPO % Spk	Rec %	Det Limit	Count	Err%
S9810033589	A	Lithium - ICP-Acid Digest	ug/g	95.60	<1.00e-02	< 10.30	<9.48e0	n/a	n/a	93.20	10.20	n/a	n/a
S9810033589	A	Manganese - ICP-Acid Digest	ug/g	92.00	<1.00e-02	< 10.30	<9.48e0	n/a	n/a	96.00	10.20	n/a	n/a
S9810033589	A	Nickel - ICP-Acid Digest	ug/g	113.8	1.180	2.30e+05	2.32e+05	2.31e+05	0.87	1.97e+03	102.0	n/a	n/a
S9810033589	A	Nickel - ICP-Acid Digest	ug/g	91.60	<2.00e-02	60.60	59.40	60.00	2.00	97.40	20.20	n/a	n/a
S9810033589	A	Silicon - ICP-Acid Digest	ug/g	163.6	2.230	2.75e+02	333.0	304.0	19.1	136.8	51.50	n/a	n/a
S9810033589	A	Silicon - ICP-Acid Digest	ug/g	92.90	<5.00e-01	<5.13e+02	<4.74e2	n/a	n/a	95.30	515.0	n/a	n/a
S9810033589	A	Zinc - ICP-Acid Digest	ug/g	83.60	1.10e-02	< 10.30	<9.48e0	n/a	n/a	96.20	10.20	n/a	n/a
S9810033589	A	Zinc - ICP-Acid Digest	ug/g	92.40	<1.00e-02	< 10.30	<9.48e0	n/a	n/a	95.40	10.20	n/a	n/a
S9810033590	M	Zirconium - ICP-Acid Digest	ug/g	100.8	<27.96	1.25e+02	126.0	125.5	0.80	100.9	1.00e-02	n/a	n/a
S9810033590	M	Zirconium (VI) by Spec.	ug/g	96.51	<24821.0	<2.4352	<2.4352	n/a	n/a	n/a	2.48e+04	n/a	n/a
S9810033590	M	Chromium - ICP-Acid Digest	ug/g	95.80	<6.20e-02	<1.35e+02	<1.35e2	n/a	n/a	95.80	135.4	n/a	n/a
S9810033590	M	Formate by IC-Dioxex 4000/4500	ug/g	100.5	<1.20e-02	1.44e+02	1.52e+03	1.48e+03	5.41	96.79	240.7	n/a	n/a
S9810033590	M	Fluoride by IC-Dioxex 4000/4500	ug/g	97.86	<1.70e-02	4.21e+02	453.0	437.2	7.32	87.56	341.0	n/a	n/a
S9810033590	M	Chloride by IC-Dioxex 4000/4500	ug/g	94.37	1.24e-01	5.51e+03	5.49e+03	5.50e+03	0.36	85.63	2.17e+03	n/a	n/a
S9810033590	M	Nitrite by IC-Dioxex 4000/4500	ug/g	95.22	<1.39e-01	<2.51e+03	<2.46e3	n/a	n/a	84.98	2.51e+03	n/a	n/a
S9810033590	M	Bromide by Ion Chromatograph	ug/g	95.95	<1.39e-01	6.39e+05	6.44e+05	6.41e+05	0.78	87.67	2.79e+03	n/a	n/a
S9810033590	M	Nitrate by IC-Dioxex 4000/4500	ug/g	98.91	<1.20e-01	4.16e+03	3.88e+03	4.02e+03	6.97	101.8	2.41e+03	n/a	n/a
S9810033590	M	Phosphate by IC-Dioxex 4000/4500	ug/g	96.39	<1.38e-01	5.43e+03	6.12e+03	5.77e+03	11.9	87.15	2.77e+03	n/a	n/a
S9810033590	M	Sulfate by IC-Dioxex 4000/4500	ug/g	104.4	<1.09e-01	3.13e+03	2.99e+03	3.06e+03	4.58	93.49	2.11e+03	n/a	n/a
S9810033590	M	Oxalate by IC-Dioxex 4000/4500	ug/g										

Table 1. Beta Summary Report.
BY-105 (2)

CORE NUMBER: 251

SEGMENT #: 1

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	RPD %	Det Limit	Count Err%
S981002955		DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	58.69	71.84	65.16	19.9	n/a	n/a	n/a
S981002955		DSC Exotherm on Perkin Elmer	Joules/g	100.8	n/a	27.66	33.76	30.71	19.9	n/a	n/a	n/a
S981002955		Ammonia by ISE-Std Additions	ppm	105.6	<10.00	47.70	41.80	44.75	13.2	105.7	10.00	n/a
S981002955		Specific Gravity	g/g	98.92	n/a	1.426	1.432	1.429	0.42	n/a	1.00e-03	n/a
S981002955		% Water by TGA on Perkin Elmer	%	98.99	n/a	53.16	52.59	52.88	1.08	n/a	9.00e-03	5.00e+02
S981002955		Alpha In Liquid Samples	uCi/mL	97.78	<1.26e-02	<4.00e-03	<5.72E-3	n/a	n/a	94.57	2.30e-04	2.07e+00
S981002955		Pur-297/240 by TRU-SPEC Resin	uCi/mL	101.0	2.77e-04	1.02e-01	1.03e-01	1.02e-01	0.98	n/a	1.15e-05	8.38E+00
S981002960		Aluminum-ICP-Acid Dil.	uCi/mL	110.7	<8.17e-06	<1.5e-05	9.72e-06	n/a	n/a	n/a	30.10	n/a
S981002960		Boron-ICP-Acid Dil.	uCi/mL	97.80	<5.00e-02	<2.67e-04	<2.67e-04	0.00	1.270e1	n/a	30.10	n/a
S981002960		Barium-ICP-Acid Dil.	uCi/mL	101.2	<5.00e-02	<30.10	<3.01e1	n/a	n/a	91.50	30.10	n/a
S981002960		Bismuth-ICP-Acid Dil.	uCi/mL	96.80	<5.00e-02	<30.10	<3.01e1	n/a	n/a	98.70	60.10	n/a
S981002960		Calcium-ICP-Acid Dil.	uCi/mL	100.2	<1.00e-01	<82.20	<6.01e1	n/a	n/a	92.90	60.10	n/a
S981002960		Chromium-ICP-Acid Dil.	uCi/mL	99.80	<1.00e-02	1.27e+02	1.28e+03	1.28e+03	0.92	n/a	30.10	n/a
S981002960		Iron-ICP-Acid Dil.	uCi/mL	96.40	<5.00e-02	32.20	31.70	31.95	1.56	78.90	300.0	n/a
S981002960		Lithium-ICP-Acid Dil.	uCi/mL	97.40	<1.00e-02	4.15e+03	4.14e+03	0.24	n/a	94.30	60.10	n/a
S981002960		Manganese-ICP-Acid Dil.	uCi/mL	99.80	<1.00e-02	<6.010	<6.01e0	n/a	n/a	94.30	60.10	n/a
S981002960		Nickel-ICP-Acid Dil.	uCi/mL	99.20	<1.00e-01	2.07e+05	2.06e+05	2.07e+05	0.48	-7.60e2	60.10	n/a
S981002960		Silicon-ICP-Acid Dil.	uCi/mL	103.8	<5.00e-02	<12.00	<1.20e1	n/a	n/a	104.0	30.10	n/a
S981002960		Sulfur-ICP-Acid Dil.	uCi/mL	105.5	<5.00e-02	1.27e+02	128.0	127.5	0.78	n/a	300.0	n/a
S981002960		Titanium-ICP-Acid Dil.	uCi/mL	96.20	<1.00e-02	<6.010	<6.01e0	n/a	n/a	94.00	60.10	n/a
S981002960		Zinc-ICP-Acid Dil.	uCi/mL	96.80	<1.00e-02	<6.010	<6.01e0	n/a	n/a	95.20	60.10	n/a
S981002960		Formate by Ion Chromatograph	uCi/mL	98.32	<2.10e-02	1.15e+03	1.14e+03	1.15e+03	0.87	110.9	37.57	n/a
S981002960		Fluoride-IC-Dioxex 4000/4500	uCi/mL	101.1	<1.20e-02	4.31e+02	4.36.0	433.6	1.15	126.9	61.81	n/a
S981002960		Chloride-IC-Dioxex 4000/4500	uCi/mL	105.2	<1.70e-02	4.38e+03	4.33e+03	4.35e+03	1.15	96.2	87.57	n/a
S981002960		Nitrite-IC-Dioxex 4000/4500	uCi/mL	102.5	<1.08e-01	6.21e+04	6.20e+04	6.20e+04	0.16	104.6	556.3	n/a
S981002960		Bromide by Ion Chromatograph	uCi/mL	99.52	<1.25e-01	<6.44e+02	<6.44e2	n/a	n/a	91.35	643.9	n/a
S981002960		Sulfate by IC-Dioxex 4000/4500	uCi/mL	99.57	<1.39e-01	2.32e+05	2.32e+05	2.32e+05	0.00	81.23	710.8	n/a
S981002960		Phosphate by IC-Dioxex 4000/4500	uCi/mL	104.0	<1.58e-01	1.84e+03	1.84e+03	1.84e+03	0.00	98.59	710.8	n/a
S981002960		Sulfate by IC-Dioxex 4000/4500	uCi/mL	106.8	<1.05e-01	5.41e+02	<5.41e2	n/a	n/a	102.6	540.9	n/a
S981002960		Oxalate-IC-Dioxex 4000/4500	uCi/mL	102.1	<1.20e-03	1.72e+03	1.71.0	172.0	1.16	n/a	1.00e-03	0.140
S981002960		Cesium-137 by GEA	uCi/mL	102.4	<2.58e-01	1.12e+03	1.12e+03	1.12e+03	0.00	99.50	1.00e-03	n/a
S981002960		Chromium (VI) by Spec.	uCi/mL	105.0	<1.20e-04	<8.60	<4.86e0	n/a	n/a	n/a	3.860	n/a
S981002965		D 1129 by ICP/MS Acid Addition	ug/mL	101.2	<9.400	3.79e+04	3.90e+04	3.84e+04	2.86	n/a	2.50e-03	n/a
S981002965		TOC by Persulfate/Coulometry	ug/mL	92.00	<1.20e-05	2.18e+03	2.16e+03	2.17e+03	0.92	98.10	40.00	n/a
S981002965		Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	n/a	4.86e-01	n/a
S981002965		Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	n/a	4.86e-01	n/a
S981002965		Uranium-235 by ICP/MS Acid Add	ug/mL	101.4	<1.20e-05	<4.86e-01	<4.86e-1	n/a	n/a	95.64	4.86e-01	n/a
S981002965		Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<1.60e-03	<6.48e-01	<6.48e-1	n/a	n/a	n/a	6.48e-01	n/a
S981002965		Uranium-238 by ICP/MS Acid Add	ug/mL	104.0	<1.20e-03	5.79e-01	5.31e-01	5.55e-01	8.65	97.50	4.86e-01	n/a
S981002965		D 1c 99 by ICP/MS Acid Addition	ug/mL	85.37	<2.00e-06	5.067	5.070	5.069	0.00	n/a	8.10e-02	n/a

26-mar-1999 07:20:02
A-0002-1

Sample#	R.#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002965	D	Sn 126 by ICP/MS	ug/mL	n/a	<1.50e-04	< 6.075	<6.09e0	n/a	n/a	n/a	6.075	n/a
S981002965	A	Neptunium-237 by ICP/MS	ug/mL	120.6	<5.74e-06	<2.32e-01	<2.32e-1	n/a	n/a	109.3	2.32e-01	n/a
S981002965	A	Plutonium-239 by ICP/MS	ug/mL	102.2	<6.51e-06	<2.64e-01	<2.64e-1	n/a	n/a	94.12	2.64e-01	n/a
S981002965	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<3.88e-06	<1.57e-01	<1.57e-1	n/a	n/a	n/a	1.57e-01	n/a
S981002965	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<3.80e-06	<1.54e-01	<1.54e-1	n/a	n/a	n/a	1.54e-01	n/a
S981002965	A	Pu/Am-241 by ICP/MS	ug/mL	108.4	<3.66e-06	<1.48e-01	<1.48e-1	n/a	n/a	99.90	1.48e-01	n/a

Table 1. Data Summary Report.
81-105 (2)

CORE NUMBER: 251

SEGMENT #: 2

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S9810022975		Bulk Density of Sample	g/mL	n/a	n/a	dry	n/a	n/a	n/a	5,00e-01	n/a
S9810022980		ISC Exotherm Dry Calculated	Joules/g	n/a	0.00e+00	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S9810022980		DSC Exotherm on Perkin Elmer	Joules/g	99.02	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a
S9810022980		% Water by TGA on Perkin Elmer	%	99.16	4.60	30.89	23.70	27.30	26.3	n/a	n/a
S9810022980		TIC by Acid/Coulometry	ug/g	96.68	2.100	5.69e+03	6.24e+03	5.86e+03	5.97	94.90	5,000
S9810022980		TOC by Persulfate/Coulometry	ug/g	95.67	2.100	2.15e+03	2.04e+03	2.20e+03	4.10	70.80	40,000
S9810022986		F Strontium-89/90 High Level	uCi/g	99.00	1.00e-02	3.810	4.430	4.120	15.0	n/a	5,00e+03
S9810022986		F Aluminum - ICP-Fusion	ug/g	99.32	5.00e-02	3.62e+04	4.47e+04	4.04e+04	21.0	93.80	1,04e+03
S9810022986		F Boron - ICP-Fusion	ug/g	97.00	5.00e-02	4.04e+03	<1.03e3	n/a	n/a	99.30	1,04e+03
S9810022986		F Barium - ICP-Fusion	ug/g	97.00	5.00e-02	4.04e+03	<1.03e3	n/a	n/a	91.80	1,04e+03
S9810022986		F Bismuth - ICP-Fusion	ug/g	101.2	4.10e-01	2.08e+03	<2.07e3	n/a	n/a	98.30	2,08e+03
S9810022986		F Calcium - ICP-Fusion	ug/g	100.4	4.10e-01	3.70e+03	2.41e+03	3.06e+03	42.2	104.0	2,08e+03
S9810022986		F Chromium - ICP-Fusion	ug/g	100.4	4.10e-02	4.84e+02	487.0	486.5	0.21	99.80	208.0
S9810022986		F Iron - ICP-Fusion	ug/g	99.80	5.00e-02	3.38e+03	2.52e+03	2.95e+03	29.2	95.00	1,04e+03
S9810022986		F Lithium - ICP-Fusion	ug/g	99.80	4.10e-02	<2.08e+02	<2.07e2	n/a	n/a	92.60	208.0
S9810022986		F Manganese - ICP-Fusion	ug/g	99.40	4.10e-02	<2.08e+02	<2.07e2	n/a	n/a	96.60	2,08e+03
S9810022986		F Nickel - ICP-Fusion	ug/g	100.0	1.39e-01	1.92e+03	1.93e+03	1.93e+03	48.1	97.80	1,04e+03
S9810022986		F Silicon - ICP-Fusion	ug/g	104.4	5.00e-02	2.95e+04	1.74e+04	2.34e+04	51.6	104.0	1,04e+03
S9810022986		F Uranium - ICP-Fusion	ug/g	96.00	4.10e-02	<2.08e+02	<2.07e2	n/a	n/a	92.30	1,04e+04
S9810022986		F Zinc - ICP-Fusion	ug/g	97.80	4.10e-02	<2.08e+02	<2.07e2	n/a	n/a	97.10	208.0
S9810022986		F Zirconium - ICP-Fusion	ug/g	104.6	3.40e-02	7.06	48.60	62.33	44.1	95.00	208.0
S9810022986		F Cesium-137 by DEA	uCi/g	96.11	3.40e-02	<1.54e-01	1.88e-01	n/a	n/a	94.57	2,05e-01
S9810022986		F Alpha of Digested Solid	uCi/g	109.0	<1.20e-03	<4.67e	<4.80e0	n/a	n/a	n/a	4,677
S9810022990		A 1129 by ICP/MS Act18/IA50159	ug/g	n/a	<1.20e-03	<4.68e-01	<4.80e-01	n/a	n/a	n/a	4,68e-01
S9810022990		A Uranium-235 by ICP/MS Act18/IA50159	ug/g	n/a	<1.20e-04	<4.68e-01	<4.80e-01	n/a	n/a	n/a	4,68e-01
S9810022990		A Uranium-238 by ICP/MS Act18/IA50159	ug/g	106.0	<1.20e-04	1.244	8.56e-01	1.050	36.6	111.8	4,68e-01
S9810022990		A Uranium-235 by ICP/MS Act18/IA50159	ug/g	104.0	<1.20e-04	<6.24e-01	<6.40e-01	n/a	n/a	n/a	4,68e-01
S9810022990		A Uranium-238 by ICP/MS Act18/IA50159	ug/g	104.0	<1.20e-04	1.77e-02	125.0	150.8	34.4	109.0	4,68e-01
S9810022990		A Technetium-99 by ICP/MS Act18/IA50159	ug/g	86.29	3.30e-03	<1.74e	1.080	1.127	8.00	84.47	7,80e-02
S9810022990		A Neptunium-237 by ICP/MS	ug/g	115.4	<1.21e-05	<4.73e-02	<4.86e-02	n/a	n/a	113.4	4,70e-02
S9810022990		A Plutonium-239 by ICP/MS	ug/g	97.30	<1.21e-05	<1.48e-01	<1.51e-01	n/a	n/a	97.61	1,48e-01
S9810022990		A Plutonium-240 by ICP/MS	ug/g	n/a	<9.50e-06	<2.79e-02	<2.88e-02	n/a	n/a	n/a	3,70e-02
S9810022990		A Plutonium-242 by ICP/MS	ug/g	101.4	<7.15e-06	<2.09e-02	<2.05e-02	n/a	n/a	102.3	2,80e-02
S9810022990		A Pu/Am-241 by ICP/MS	ug/g	93.80	9.10e-05	2.61e-04	2.50e-04	2.30e-04	26.5	n/a	5,00e-02
S9810022990		A Aluminum - ICP-Acid Digest	ug/g	100.4	4.63e-01	55.50	53.50	54.50	3.67	88.20	38.80
S9810022990		A Boron - ICP-Acid Digest	ug/g	95.20	5.00e-02	<38.80	<39.861	n/a	n/a	86.80	38.80
S9810022990		A Bismuth - ICP-Acid Digest	ug/g	102.0	1.59e-01	<77.60	<79.661	n/a	n/a	81.00	77.60
S9810022990		A Calcium - ICP-Acid Digest	ug/g	94.00	<1.00e-02	6.20e-02	4.990	3.597	21.0	81.00	7,760
S9810022990		A Chromium - ICP-Acid Digest	ug/g	92.80	5.10e-02	2.09e+03	3.04e+03	2.56e+03	57.0	23.40	38.80
S9810022990		A Iron - ICP-Acid Digest	ug/g								

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002956	A	Potassium -ICP-Acid Digest	ug/g	86.60	< 5.00e-01	1.43e+03	1.10e+03	1.26e+03	26.1	52.80	588.0
S98T002956	A	Lithium -ICP-Acid Digest	ug/g	93.40	< 1.00e-02	< 7.60	< 7.90	64.95	n/a	83.40	n/a
S98T002956	A	Manganese -ICP-Acid Digest	ug/g	93.00	< 1.00e-02	67.40	62.50	64.95	n/a	7.760	n/a
S98T002956	A	Sodium -ICP-Acid Digest	ug/g	117.4	7.45e-01	1.81e+05	1.82e+05	1.82e+05	0.55	-1.47083	7.60
S98T002956	A	Nickel -ICP-Acid Digest	ug/g	92.40	< 2.00e-02	64.00	60.20	62.10	6.12	98.20	n/a
S98T002956	A	Silicon -ICP-Acid Digest	ug/g	150.0	7.59e-01	50.60	61.40	52.33	17.0	33.00	n/a
S98T002956	A	Uranium -ICP-Acid Digest	ug/g	90.90	< 5.00e-01	< 3.88e-02	< 3.98e-02	12.80	n/a	388.0	n/a
S98T002956	A	Zinc -ICP-Acid Digest	ug/g	86.20	< 3.00e-02	72.30	72.80	12.35	3.96	7.760	n/a
S98T002956	A	Zirconium -ICP-Acid Digest	ug/g	92.80	< 6.00e-02	< 7.60	< 7.90	249.7	11.6	98.20	n/a
S98T002956	A	Formate by IC-Dioxin 4000/4500	ug/g	94.53	< 2.00e-02	2.68e+02	2.35e+02	2.097	1.6	72.66	n/a
S98T002956	A	Fluoride-IC-Dioxin 4000/4500	ug/g	102.4	< 1.20e-02	2.08e+03	2.08e+03	2.08e+03	5.44	100.5	n/a
S98T002956	A	Chloride-IC-Dioxin 4000/4500	ug/g	102.6	< 1.00e-02	1.08e+03	1.08e+03	1.08e+03	2.44	101.0	n/a
S98T002956	A	Nitrite-IC - Dioxin 4000/4500	ug/g	102.4	< 5.20e-01	1.56e+04	1.29e+04	1.35e+04	5.28	101.2	n/a
S98T002956	A	Bromide by Ion Chromatograph	ug/g	100.3	< 1.00e-01	1.24e+03	1.24e+03	1.24e+03	92.66	1.24e+03	n/a
S98T002956	A	Nitrate by IC-Dioxin 4000/4500	ug/g	102.7	< 1.70e-01	4.24e+03	4.08e+03	4.51e+03	12.6	118.8	n/a
S98T002956	A	Phosphate-IC-Dioxin 4000/4500	ug/g	104.9	< 1.20e-01	1.59e+03	1.59e+03	2.30e+03	10.8	100.2	n/a
S98T002956	A	Sulfate by IC-Dioxin 4000/4500	ug/g	102.8	< 1.58e-01	2.78e+03	2.85e+03	3.72e+03	2.27	99.37	n/a
S98T002956	A	Oxalate-IC-Dioxin 4000/4500	ug/g	106.3	< 1.05e-01	2.88e+03	2.82e+03	2.72e+03	5.09	100.7	n/a

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002956	A	DSC Exotherm Dry	Joules/g Dry	100.8	n/a	63.22	61.51	61.86	1.15	n/a	n/a
S98T002956	A	DSC Exotherm on Perkin Elmer	Joules/g	100.8	n/a	29.50	29.16	29.33	1.16	n/a	n/a
S98T002956	A	Ammonia by ISE-Std Additions	ug/mL	103.3	< 10.00	< 10.00	< 10.00	1.433	n/a	10.00	n/a
S98T002956	A	Specific Gravity	99.99	98.99	< 1.440	1.440	1.437	1.437	0.49	1.00e-03	n/a
S98T002956	A	% Water by IGA on Perkin Elmer	%	98.99	n/a	52.80	52.38	52.59	0.80	n/a	n/a
S98T002956	A	Alpha in Liquid Samples	ug/mL	94.44	< 5.75e-03	< 9.42e-03	< 9.73e-3	7.62e-02	3.94	1.10e-02	3.40e+02
S98T002961	D	Strontium-89/90 High Level	ug/mL	101.0	< 2.77e-04	< 7.47e-04	< 7.77e-04	7.62e-02	3.94	2.25e-04	2.40e+00
S98T002961	D	Pu-239/240 by TRU-SPEC resin	ug/mL	101.0	< 8.17e-04	< 9.98e-06	< 8.53e-06	2.92e+04	2.37	7.98e-06	8.01e+00
S98T002961	D	Aluminum-ICP-Acid Dtl.	ug/mL	97.40	< 5.00e-02	< 2.99e+04	< 3.01e+04	2.96e+04	2.37	30.10	n/a
S98T002961	D	Boron-ICP-Acid Dtl.	ug/mL	96.60	< 5.00e-02	< 30.10	< 30.10	n/a	n/a	30.10	n/a
S98T002961	D	Barium-ICP-Acid Dtl.	ug/mL	99.60	< 1.00e-01	< 30.10	< 30.10	n/a	n/a	30.10	n/a
S98T002961	D	Bismuth-ICP-Acid Dtl.	ug/mL	101.4	< 1.00e-01	< 60.10	< 60.10	n/a	n/a	60.10	n/a
S98T002961	D	Calcium-ICP-Acid Dtl.	ug/mL	98.80	< 1.00e-02	< 1.24e+03	< 1.24e+03	3.28	n/a	60.10	n/a
S98T002961	D	Chromium-ICP-Acid Dtl.	ug/mL	95.60	< 5.00e-02	< 34.70	< 32.70	33.70	5.93	300.0	n/a
S98T002961	D	Cobalt-ICP-Acid Dtl.	ug/mL	95.80	< 5.00e-01	< 4.23e+03	< 3.97e+03	4.10e+03	6.34	300.0	n/a
S98T002961	D	Copper-ICP-Acid Dtl.	ug/mL	101.6	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	6.010	n/a
S98T002961	D	Lithium-ICP-Acid Dtl.	ug/mL	97.40	< 1.00e-02	< 2.04e+05	< 2.00e+05	2.02e+05	1.98	60.10	n/a
S98T002961	D	Selenium-ICP-Acid Dtl.	ug/mL	100.8	< 2.00e-02	< 12.00	< 12.00	141.0	5.67	12.00	n/a
S98T002961	D	Nickel-ICP-Acid Dtl.	ug/mL	98.60	< 5.00e-02	< 1.45e+02	< 1.37e+02	n/a	n/a	300.10	n/a
S98T002961	D	Silicon-ICP-Acid Dtl.	ug/mL	102.6	< 5.00e-01	< 3.01e+02	< 3.01e+02	n/a	n/a	300.10	n/a
S98T002961	D	Uranium-ICP-Acid Dtl.	ug/mL	97.20	< 5.00e-02	< 6.010	< 6.010	n/a	n/a	6.010	n/a
S98T002961	D	Zinc-ICP-Acid Dtl.	ug/mL	93.40	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	6.010	n/a
S98T002961	D	Zirconium-ICP-Acid Dtl.	ug/mL	96.80	< 1.00e-02	< 6.44e+02	< 652.0	647.8	1.23	n/a	n/a
S98T002961	D	Formate by Ion Chromatograph	ug/mL	98.32	< 2.10e-02	6.44e+02	6.44e+02	409.7	0.24	61.81	n/a
S98T002961	D	Fluoride-IC-Dioxin 4000/4500	ug/mL	101.1	< 1.20e-02	4.09e+02	4.10e+02	4.24e+03	1.17	n/a	n/a
S98T002961	D	Chloride-IC-Dioxin 4000/4500	ug/mL	105.2	< 1.70e-02	6.24e+03	6.24e+03	6.24e+03	2.76	556.3	n/a
S98T002961	D	Nitrite-IC - Dioxin 4000/4500	ug/mL	102.5	< 1.08e-01	6.07e+04	6.24e+04	6.15e+04	2.76	n/a	n/a
S98T002961	D	Bromide by Ion Chromatograph	ug/mL	99.52	< 1.25e-01	< 6.44e+02	< 6.44e+02	n/a	n/a	643.9	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002961		Nitrate by IC-Dionex 4000/4500	ug/mL	97.7	<1.30e-01	2.08e-05	2.09e-05	2.09e-05	0.48	N/A	716.0
S981002961		Phosphate by IC-Dionex 4000/4500	ug/mL	104.0	<1.20e-01	1.38e-02	1.38e-02	1.38e-02	6.75	N/A	618.1
S981002961		Sulfate by IC-Dionex 4000/4500	ug/mL	104.4	<1.50e-01	5.41e-02	1.80e-03	1.74e-03	6.90	N/A	710.8
S981002961		Oxalate by IC-Dionex 4000/4500	ug/mL	104.8	<1.05e-01	5.41e-02	<5.11e-02	<5.11e-02	N/A	N/A	540.9
S981002961		Cesium-137 by GEA	ug/mL	102.4	<1.00e-03	1.81e-02	18.46	183.5	0.52	N/A	0.130
S981002966		Chromium (VI) by Spec.	ug/mL	102.4	<2.58e-03	1.10e-02	1.09e-03	1.10e-03	0.91	N/A	1.00e-02
S981002966	D	112 by IC/MS Acid Addition	ug/mL	105.0	<1.20e-04	<1.860	<2.860	<2.860	N/A	N/A	4.860
S981002966		OH- by Pot. Titration	ug/mL	101.2	<1.20e-04	3.88e-02	4.00e-04	3.94e-04	3.75	N/A	5.00e-03
S981002966		TOL by Persulfate/Oxidometry	ug/mL	95.00	3.300	2.28e-03	2.17e-03	2.22e-03	4.94	N/A	40.00
S981002966	D	Uranium-235 by IC/MS Acid Add	ug/mL	N/A	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	N/A	N/A	4.86e-01
S981002966	D	Uranium-235 by IC/MS Acid Add	ug/mL	107.2	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	N/A	N/A	4.86e-01
S981002966	D	Uranium-235 by IC/MS Acid Add	ug/mL	107.2	<1.20e-05	<4.86e-01	<4.86e-01	<4.86e-01	N/A	N/A	4.86e-01
S981002966	D	Uranium-238 by IC/MS Acid Add	ug/mL	112.0	<1.20e-05	5.47e-01	5.43e-01	5.45e-01	0.73	N/A	4.86e-01
S981002966	D	Uranium-238 by IC/MS Acid Add	ug/mL	92.68	<2.00e-04	5.47e-01	5.43e-01	5.45e-01	0.73	N/A	4.86e-01
S981002966	D	IC 90 by IC/MS Acid Addition	ug/mL	102.4	<1.50e-04	<3.045	<4.830	<4.888	2.45	N/A	4.10e-02
S981002966	D	Sn 126 by IC/MS Acid Addition	ug/mL	102.6	<5.76e-04	<2.32e-01	<2.32e-01	<2.32e-01	N/A	N/A	3.045
S981002966	A	Plutonium-239 by IC/MS	ug/mL	102.2	<6.51e-04	<2.64e-01	<2.64e-01	<2.64e-01	N/A	N/A	2.32e-01
S981002966	A	Plutonium-240 by IC/MS	ug/mL	102.2	<3.88e-04	<1.57e-01	<1.57e-01	<1.57e-01	N/A	N/A	1.57e-01
S981002966	A	Plutonium-242 by IC/MS	ug/mL	N/A	<3.80e-04	<1.54e-01	<1.54e-01	<1.54e-01	N/A	N/A	1.54e-01
S981002966	A	Pu/Am-241 by IC/MS	ug/mL	108.4	<3.68e-04	<1.48e-01	<1.48e-01	<1.48e-01	N/A	N/A	1.48e-01

Table 1. Data Summary Report.
81-105 (2)CORE NUMBER: 251
SEGMENT #: 3

SEGMENT PORTION: Drainable Liquid

Sample#	RAW Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002957	DSC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	9.690	30.87	20.48	n/a	n/a	n/a	n/a
S981002957	DSC Exotherm Wet Partition Elmer	Joules/g	97.19	n/a	4.990	15.89	10.44	104	n/a	n/a	n/a
S981002957	Amonia by TS/SC Additions	ppb	102.3	<10.00	<10.00	<1.00E+1	n/a	100.0	n/a	10.00	n/a
S981002957	Specific Gravity	Sp.G.	98.22	n/a	1.484	1.438	1.461	3.15	n/a	1.00E-03	n/a
S981002957	% Water by Tit on Perkin Elmer	%	95.26	n/a	57.98	49.08	48.53	2.27	n/a	n/a	n/a
S981002957	Alpha in Liquid Samples	UG/mL	101.4	<5.7E-02	1.55E-02	<1.28E-2	n/a	n/a	n/a	n/a	n/a
S981002962	Strentium-89/90 High Level	UG/mL	101.4	7.7E-04	8.19E-02	9.06E-02	8.97E-02	1.89	n/a	2.20E-02	1.7E+02
S981002962	Pu-239/240 by TRU-SPEC Resin	UG/mL	110.9	<5.7E-02	1.13E-02	1.14E-02	1.12E-02	2.67	n/a	2.29E-04	2.17E+00
S981002962	Boregon-10CP-Acid Dill.	UG/mL	97.40	<5.00E-02	3.87E-04	3.65E-04	3.74E-04	5.07	46.60	30.10	n/a
S981002962	Boregon-10CP-Acid Dill.	UG/mL	102.3	<5.00E-02	<30.10	3.01E1	n/a	n/a	104.0	30.10	n/a
S981002962	Barium-137CP-Acid Dill.	UG/mL	96.60	<5.00E-02	<60.10	<6.01E1	n/a	n/a	92.40	30.10	n/a
S981002962	Barium-137CP-Acid Dill.	UG/mL	99.60	<1.00E-01	<60.10	<6.01E1	n/a	n/a	92.40	30.10	n/a
S981002962	Calcium-137CP-Acid Dill.	UG/mL	101.4	<1.00E-01	<60.10	<6.01E1	n/a	n/a	92.40	30.10	n/a
S981002962	Chromium-10CP-Acid Dill.	UG/mL	98.80	<1.00E-02	1.23E+03	1.16E+03	1.20E+03	5.78	103.50	60.10	n/a
S981002962	Iron-59CP-Acid Dill.	UG/mL	95.60	<1.00E-02	36.50	36.80	37.45	2.55	92.40	30.10	n/a
S981002962	Potassium-10CP-Acid Dill.	UG/mL	95.80	<1.00E-01	4.51E+03	4.21E+03	4.36E+03	6.86	82.40	30.10	n/a
S981002962	Lithium-10CP-Acid Dill.	UG/mL	101.6	<1.00E-02	<6.010	<6.01E0	n/a	n/a	92.40	30.10	n/a
S981002962	Manganese-10CP-Acid Dill.	UG/mL	97.40	<1.00E-02	<6.010	<6.01E0	n/a	n/a	92.40	30.10	n/a
S981002962	Sodium-10CP-Acid Dill.	UG/mL	100.8	<1.00E-01	2.08E+05	1.98E+05	2.03E+05	4.93	-1.03E+2	60.10	n/a
S981002962	Nickel-10CP-Acid Dill.	UG/mL	98.60	<2.00E-02	<12.00	<1.20E1	n/a	n/a	98.00	12.00	n/a
S981002962	Silicon-10CP-Acid Dill.	UG/mL	102.6	<5.00E-02	<1.79E+02	163.0	171.0	9.36	100.0	30.10	n/a
S981002962	Uranium-10CP-Acid Dill.	UG/mL	97.20	<5.00E-02	<5.01E+02	<3.01E2	n/a	n/a	94.50	300.0	n/a
S981002962	Zinc-10CP-Acid Dill.	UG/mL	93.40	<1.00E-02	<6.010	<6.01E0	n/a	n/a	93.80	60.10	n/a
S981002962	Zirconium-10CP-Acid Dill.	UG/mL	105.9	<2.00E-02	1.35E+03	1.39E+03	1.37E+03	2.92	124.4	68.88	n/a
S981002962	Formate by Ion Chromatograph	UG/mL	105.9	<2.00E-02	3.21E+02	314.0	317.4	2.20	93.22	122.4	n/a
S981002962	Fluoride-10CP-Acid Dill.	UG/mL	100.4	<1.00E-02	4.58E+03	4.67E+03	4.63E+03	1.95	102.8	173.4	n/a
S981002962	Chloride-10CP-Acid Dill.	UG/mL	105.4	<1.90E-02	6.32E+04	6.46E+04	6.39E+04	2.19	111.6	1.10E+03	n/a
S981002962	Nitrite-10CP-Acid Dill.	UG/mL	98.89	<1.28E-01	1.80E+05	1.82E+05	1.81E+05	n/a	95.39	1.28E+03	n/a
S981002962	Nitrate by Ion Chromatograph	UG/mL	102.0	<1.39E-01	<1.22E+03	<1.22E3	n/a	n/a	102.0	1.42E+03	n/a
S981002962	Phosphate-10CP-Acid Dill.	UG/mL	102.7	<1.20E-01	<1.22E+03	<1.22E3	n/a	n/a	102.0	1.42E+03	n/a
S981002962	Sulfate by IC-Dionex 4000/4500	UG/mL	100.0	<1.38E-01	<1.07E+03	<1.07E3	2.21E+03	6.33	101.4	1.41E+03	n/a
S981002962	Oxalate-10CP-Acid Dill.	UG/mL	103.3	<1.05E-01	<1.07E+03	<1.07E3	n/a	n/a	103.5	1.07E+03	n/a
S981002962	Cesium-137 by Spec.	UG/mL	102.1	1.00E-03	2.05E+02	205.0	205.0	0.00	n/a	n/a	0.130
S981002967	Chromium (VI) by Spec.	UG/mL	101.4	<2.58E-01	1.20E+03	1.19E+03	1.20E+03	0.84	99.10	1.00E-02	n/a
S981002967	OH- by Pot. Titration	UG/mL	105.0	<1.20E-04	<4.860	<4.86E0	n/a	n/a	n/a	4.860	n/a
S981002967	TOC by Persulfate/Coulometry	UG/mL	101.2	n/a	4.89E+04	4.93E+04	4.91E+04	0.81	n/a	5.00E+03	n/a
S981002967	Uranium-233 by ICP/MS Acid Add	UG/mL	93.33	13.00	2.68E+03	2.72E+03	2.70E+03	1.48	n/a	40.00	n/a
S981002967	Uranium-234 by ICP/MS Acid Add	UG/mL	n/a	<1.20E-05	<4.86E-01	<4.86E-1	n/a	n/a	n/a	4.86E-01	n/a
S981002967	Uranium-235 by ICP/MS Acid Add	UG/mL	n/a	<1.20E-05	<4.86E-01	<4.86E-1	n/a	n/a	n/a	4.86E-01	n/a
S981002967	Uranium-236 by ICP/MS Acid Add	UG/mL	107.7	<1.20E-05	<4.86E-01	<4.86E-1	n/a	n/a	102.1	4.86E-01	n/a
S981002967	Uranium-238 by ICP/MS Acid Add	UG/mL	n/a	<1.20E-05	<4.86E-01	<4.86E-1	n/a	n/a	n/a	4.86E-01	n/a
S981002967	Uranium-238 by ICP/MS Acid Add	UG/mL	112.0	<1.20E-05	5.72E-01	5.71E-01	5.47E-01	8.97	105.5	4.86E-01	n/a
S981002967	Tc 99 by ICP/MS Acid Addition	UG/mL	92.68	<2.00E-06	6.117	6.330	6.223	3.37	97.72	4.10E-02	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002967	D	Sn 126 by ICP/MS	ug/mL	n/a	<1.50e-04	< 3.0e-5	<3.05e0	n/a	n/a	n/a	5.005	n/a
S98T002967	A	Neptunium-237 by ICP/MS	ug/mL	120.6	<5.74e-06	< 3.2e-01	<2.32e-1	n/a	n/a	114.9	2.52e-01	n/a
S98T002967	A	Plutonium-239 by ICP/MS	ug/mL	102.2	<6.51e-06	<2.64e-01	<2.64e-1	n/a	n/a	97.61	2.64e-01	n/a
S98T002967	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<3.88e-06	<1.57e-01	<1.57e-1	n/a	n/a	n/a	1.57e-01	n/a
S98T002967	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<3.80e-06	<1.54e-01	<1.54e-1	n/a	n/a	n/a	1.54e-01	n/a
S98T002967	A	Pu/Am-241 by ICP/MS	ug/mL	108.4	<3.66e-06	<1.48e-01	<1.48e-1	n/a	n/a	102.8	1.48e-01	n/a

CODE NUMBER: 251
SEGMENT #: 4

Table 1. Data Summary Report.
81-105 (2)

SEGMENT POSITION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	% Spk Rec.	Det Limit	Count Err%
S98T020297		Bulk Density of Sample	g/mL	n/a	n/a	1.920	n/a	n/a	n/a	n/a	5.00e-01	n/a
S98T020298		USC Exotherm Dry Calculated	cal/sec/g DRY	n/a	n/a	1.09e+02	29.90	69.40	116	n/a	n/a	n/a
S98T020298		USC Exotherm on Peak Elmer	cal/sec/g	99.23	n/a	25.65	17.96	41.81	114	n/a	n/a	n/a
S98T020298		% H ₂ O by TGA in Perkin Elmer	%	99.19	n/a	34.57	45.28	39.92	26.8	n/a	n/a	n/a
S98T020298		TIC by Fed/Cal Counter in Elmer	uV/g	96.48	4.60	6.82e+03	1.19e+04	9.36e+03	54.3	24.0	5.000	n/a
S98T020298		TOC by Persulfate/Conductivity	uS/g	93.67	2.100	2.86e+03	3.40e+03	3.13e+03	17.3	101.0	40.00	1.20e+00
S98T020298		F STCOTM-88/90 High Level	uS/g	100.0	1.00e-02	6.200	6.680	6.440	7.45	n/a	5.00e-03	1.20e+00
S98T020298		F Aluminum - ICP-Fusion	uS/g	98.40	5.00e-02	3.64e+04	4.39e+04	4.02e+04	18.7	99.00	1.01e+03	n/a
S98T020298		F Boron - ICP-Fusion	uS/g	103.2	5.00e-02	1.01e+03	1.01e+03	n/a	n/a	100.0	1.01e+03	n/a
S98T020298		F Barium - ICP-Fusion	uS/g	97.40	5.00e-02	1.01e+03	1.01e+03	n/a	n/a	92.40	1.01e+03	n/a
S98T020298		F Bismuth - ICP-Fusion	uS/g	101.2	1.00e-01	2.02e+03	2.02e+03	n/a	n/a	99.90	2.02e+03	n/a
S98T020298		F Calcium - ICP-Fusion	uS/g	104.8	1.00e-01	2.36e+03	2.02e+03	n/a	n/a	105.0	2.02e+03	n/a
S98T020298		F Chromium - ICP-Fusion	uS/g	100.6	1.00e-02	6.56e+02	680.0	668.0	3.59	100.0	202.0	n/a
S98T020298		F Iron - ICP-Fusion	uS/g	97.00	5.00e-02	8.36e+03	3.57e+03	5.96e+03	80.3	95.70	1.01e+03	n/a
S98T020298		F Lithium - ICP-Fusion	uS/g	99.80	1.00e-02	2.02e+02	2.02e+02	n/a	n/a	93.60	202.0	n/a
S98T020298		F Manganese - ICP-Fusion	uS/g	99.60	1.00e-01	2.02e+02	2.02e+02	n/a	n/a	97.20	202.0	n/a
S98T020298		F Sodium - ICP-Fusion	uS/g	99.40	1.00e-01	1.97e+03	1.99e+03	1.98e+03	1.01	109.0	2.02e+03	n/a
S98T020298		F Nickel - ICP-Fusion	uS/g	100.4	1.39e-01	5.08e+03	4.35e+03	4.72e+03	15.5	101.0	405.0	n/a
S98T020298		F Silicon - ICP-Fusion	uS/g	100.0	5.00e-02	7.34e+03	7.59e+03	7.46e+03	3.35	106.0	1.01e+04	n/a
S98T020298		F Uranium - ICP-Fusion	uS/g	97.30	5.00e-01	1.01e+04	1.01e+04	n/a	n/a	94.00	1.01e+04	n/a
S98T020298		F Zinc - ICP-Fusion	uS/g	96.00	1.00e-02	2.02e+02	2.02e+02	n/a	n/a	97.50	202.0	n/a
S98T020298		F Zirconium - ICP-Fusion	uS/g	105.9	1.00e-02	2.02e+02	2.02e+02	n/a	n/a	95.20	202.0	n/a
S98T020298		F Cerium-137 by GEA	uS/g	105.9	4.30e-02	65.61	62.50	64.06	4.84	n/a	n/a	0.360
S98T020298		F Alpha of Digest Solid	uS/g	96.11	2.00e-01	46.62e-02	1.84e-01	n/a	n/a	95.80	1.99e-01	5.00e+02
S98T020298		A 1129 by ICP/MS Acid/dig LA-505159	uS/g	109.0	1.20e-03	4.782	4.90e0	n/a	n/a	n/a	4.782	n/a
S98T020299		A Uranium-233 by ICP/MS Acid/dig159	uS/g	n/a	1.20e-04	44.78e-01	44.90e-1	n/a	n/a	n/a	4.78e-01	n/a
S98T020299		A Uranium-234 by ICP/MS Acid/dig159	uS/g	n/a	1.20e-04	44.78e-01	44.90e-1	n/a	n/a	n/a	4.78e-01	n/a
S98T020299		A Uranium-235 by ICP/MS Acid/dig159	uS/g	106.0	1.20e-04	1.598	1.630	1.614	1.86	113.2	4.78e-01	n/a
S98T020299		A Uranium-236 by ICP/MS Acid/dig159	uS/g	n/a	1.20e-04	46.38e-01	46.53e-1	n/a	n/a	n/a	6.38e-01	n/a
S98T020299		A Uranium-238 by ICP/MS Acid/dig159	uS/g	104.0	1.20e-04	2.44e+02	234.0	239.2	4.18	108.0	4.78e-01	n/a
S98T020299		A Tc99 by ICP/MS Acid/dig LA-505159	uS/g	86.29	3.30e-05	1.211	1.140	1.175	5.96	85.84	8.00e-02	n/a
S98T020299		A Sm126 ICP/MS Acid/dig LA-505159	uS/g	n/a	1.50e-03	4.5977	4.612e0	n/a	n/a	n/a	5.978	n/a
S98T020299		A Neptunium-237 by ICP/MS	uS/g	115.4	1.20e-02	44.94e-02	44.94e-2	n/a	n/a	110.3	4.80e-02	n/a
S98T020299		A Plutonium-239 by ICP/MS	uS/g	97.30	3.79e-05	2.08e-01	1.53e-1	n/a	n/a	94.62	1.51e-01	n/a
S98T020299		A Plutonium-240 by ICP/MS	uS/g	n/a	5.00e-06	3.79e-02	3.88e-2	n/a	n/a	n/a	3.80e-02	n/a
S98T020299		A Plutonium-242 by ICP/MS	uS/g	n/a	1.41e-05	2.85e-02	2.92e-2	n/a	n/a	n/a	2.80e-02	n/a
S98T020299		A Pu/Am-241 by ICP/MS	uS/g	101.4	1.41e-05	5.61e-02	5.78e-2	n/a	n/a	99.42	5.60e-02	n/a
S98T020299		A Aluminum - ICP-Acid Digest	uS/g	93.80	9.10e-02	3.66e+04	3.25e+04	3.46e+04	21.9	-8.80e00	39.60	n/a
S98T020299		A Boron - ICP-Acid Digest	uS/g	100.4	4.63e-01	61.30	82.50	71.90	29.9	89.00	39.60	n/a
S98T020299		A Barium - ICP-Acid Digest	uS/g	95.20	5.00e-02	3.92	4.06e1	n/a	n/a	92.00	39.60	n/a
S98T020299		A Bismuth - ICP-Acid Digest	uS/g	90.60	1.00e-01	7.930	4.812e1	n/a	n/a	91.20	79.20	n/a
S98T020299		A Calcium - ICP-Acid Digest	uS/g	102.0	1.59e-03	1.51e+03	880.0	1.20e+03	52.7	95.40	79.20	n/a
S98T020299		A Chromium - ICP-Acid Digest	uS/g	94.00	1.00e-02	5.64e+02	480.0	526.0	14.4	88.20	7.920	n/a
S98T020299		A Iron - ICP-Acid Digest	uS/g	92.80	5.10e-02	1.51e+03	1.05e+03	1.28e+03	35.9	39.60	39.60	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S9810022991	A	Potassium - ICP-Acid Digest	ug/g	86.60	<5.00e-01	1.76e+03	1.07e+03	1.09e+03	12.8	50.60	395.0
S9810022991	A	Lithium - ICP-Acid Digest	ug/g	93.40	<1.00e-02	7.930	<8.12e0	n/a	88.80	90.60	7.920
S9810022991	A	Manganese - ICP-Acid Digest	ug/g	93.00	<1.00e-02	49.40	33.80	41.60	37.5	89.20	7.920
S9810022991	A	Sodium - ICP-Acid Digest	ug/g	117.4	7.45e-01	1.70e+05	1.94e+05	1.85e+05	14.2	36.20	79.20
S9810022991	A	Nickel - ICP-Acid Digest	ug/g	92.40	<2.00e-02	80.10	72.30	76.20	10.2	92.40	15.90
S9810022991	A	Silicon - ICP-Acid Digest	ug/g	150.0	7.50e-01	1.28e+05	1.17e+05	1.22e+05	7.41	137.8	39.60
S9810022991	A	Strontium - ICP-Acid Digest	ug/g	90.90	<5.00e-01	<3.97e+02	44.00e2	n/a	98.40	396.0	n/a
S9810022991	A	Zinc - ICP-Acid Digest	ug/g	86.20	3.50e-02	7.930	8.140	9.620	24.5	90.00	7.920
S9810022991	A	Zirconium - ICP-Acid Digest	ug/g	92.80	<6.00e-02	7.930	8.480	n/a	90.60	7.920	n/a
S9810022995	W	Formate by IC-Dioxex 4000/4500	ug/g	90.30	<6.00e-02	2.93e+02	320.0	308.6	8.61	80.17	73.32
S9810022995	W	Fluoride-IC-Dioxex 4000/4500	ug/g	109.3	<1.20e-02	7.00e+03	6.99e+03	8.84e+03	88.5	112.2	97.01
S9810022995	W	Chloride-IC-Dioxex 4000/4500	ug/g	102.4	1.70e-02	1.31e+03	2.29e+03	1.29e+03	3.89	106.9	171.0
S9810022995	W	Nitrite-IC-Dioxex 4000/4500	ug/g	102.4	5.23e-01	1.65e+02	1.65e+02	1.72e+04	8.14	107.8	1.09e+03
S9810022995	W	Bromide by Ion Chromatograph	ug/g	100.3	<1.25e-01	1.22e+03	1.22e+03	n/a	n/a	96.59	2.77e+03
S9810022995	W	Nitrate by IC-Dioxex 4000/4500	ug/g	106.4	1.35e-01	2.48e+05	1.34e+05	2.42e+05	89.5	98.99	2.77e+03
S9810022995	W	Phosphate-IC-Dioxex 4000/4500	ug/g	99.43	<1.20e-01	7.48e+03	6.40e+03	5.04e+03	96.8	94.88	970.1
S9810022995	W	Sulfate by IC-Dioxex 4000/4500	ug/g	104.3	<1.38e-01	6.45e+03	2.84e+03	4.75e+03	79.7	103.8	1.12e+03
S9810022995	W	Oxalate-IC-Dioxex 4000/4500	ug/g	104.8	<1.05e-01	5.52e+03	1.05e+04	8.01e+03	62.0	103.5	848.7

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S9810022958		DSC Exotherm Dry Calculated	Joules/g Dry			21.77	17.00	19.38	24.6	n/a	n/a
S9810022958		DSC Exotherm on Perkin Elmer	Joules/g	99.10	n/a	11.76	9.180	10.47	24.6	n/a	n/a
S9810022958		Ammonia by ISE-Std Additions	ug/mL	98.12	<10.00	<10.00	<1.00E+1	n/a	95.28	10.00	n/a
S9810022958		Specific Gravity	g/g	98.99	n/a	1.465	1.447	1.456	1.24	n/a	1.00e-03
S9810022958		% Water by IGA on Perkin Elmer	%	98.86	n/a	66.51	45.47	45.99	2.26	n/a	n/a
S9810022958		Alpha In Liquid Samples	ug/mL	94.64	<5.79e-03	<2.17e-02	1.43e-02	n/a	n/a	2.20e-02	2.14E+02
S9810022958		Strontium-89/90 High Level	ug/mL	110.1	2.77e-04	9.36e-02	9.41e-02	9.39e-02	0.53	n/a	2.25e-04
S9810022963	D	PU-239/240 by TRU-SPEC Resin	ug/mL	110.7	<8.17e-06	<7.90e-06	<7.84E-6	n/a	n/a	7.90e-06	9.08E+00
S9810022963	D	Aluminum-IC-Acid Dtl.	ug/mL	97.80	<5.00e-02	4.72e+04	4.40e+04	4.56e+04	7.02	<4.62e02	30.10
S9810022963	D	Boron-IC-Acid Dtl.	ug/mL	102.4	<5.00e-02	34.60	31.30	32.95	10.0	99.10	30.10
S9810022963	D	Boron-IC-Acid Dtl.	ug/mL	97.00	<5.00e-02	<30.10	<6.01e1	n/a	n/a	92.70	30.10
S9810022963	D	Barium-IC-Acid Dtl.	ug/mL	100.4	<1.00e-01	<60.10	<6.01e1	n/a	n/a	95.30	60.10
S9810022963	D	Bismuth-IC-Acid Dtl.	ug/mL	103.2	<1.00e-01	<60.10	<6.01e1	n/a	n/a	104.0	60.10
S9810022963	D	Calcium-IC-Acid Dtl.	ug/mL	99.80	<1.00e-02	1.27e+03	1.18e+03	1.22e+03	7.35	82.30	6.010
S9810022963	D	Chromium-IC-Acid Dtl.	ug/mL	96.40	<5.00e-02	44.30	40.70	42.50	8.47	92.80	30.10
S9810022963	D	Iron-IC-Acid Dtl.	ug/mL	94.40	<5.00e-02	4.79e+03	4.44e+03	4.62e+03	7.58	45.10	300.0
S9810022963	D	Potassium-IC-Acid Dtl.	ug/mL	99.80	<1.00e-02	<6.010	<6.01e0	n/a	n/a	94.90	6.010
S9810022963	D	Silicon-IC-Acid Dtl.	ug/mL	98.80	<1.00e-01	2.11e+05	1.96e+05	2.04e+05	7.37	<2.52e03	60.10
S9810022963	D	Sodium-IC-Acid Dtl.	ug/mL	98.80	<2.00e-02	<12.00	<1.20e1	n/a	96.30	12.00	n/a
S9810022963	D	Strontium-IC-Acid Dtl.	ug/mL	104.0	<5.00e-02	2.73e+02	255.0	264.0	6.82	101.0	300.10
S9810022963	D	Tellurium-IC-Acid Dtl.	ug/mL	96.50	<5.00e-02	<3.01e02	<3.01e02	n/a	n/a	94.00	300.10
S9810022963	D	Zinc-IC-Acid Dtl.	ug/mL	94.80	<1.00e-02	<6.010	<6.01e0	n/a	n/a	95.70	6.010
S9810022963	D	Zirconium-IC-Acid Dtl.	ug/mL	97.20	<1.00e-02	<6.010	<6.01e0	n/a	n/a	110.7	37.57
S9810022963	D	Formate by Ion Chromatograph	ug/mL	104.2	<4.30e-02	1.42e+03	1.42e+03	n/a	0.00	122.7	122.4
S9810022963	D	Fluoride-IC-Dioxex 4000/4500	ug/mL	100.8	<1.20e-02	<1.22e+02	<1.22e+02	n/a	n/a	99.63	173.4
S9810022963	D	Chloride-IC-Dioxex 4000/4500	ug/mL	105.4	<1.00e-02	5.30e+03	5.12e+03	5.21e+03	3.45	107.0	1.10e+03
S9810022963	D	Nitrite-IC-Dioxex 4000/4500	ug/mL	105.2	<1.00e-02	7.31e+04	7.09e+04	7.20e+04	3.06	107.0	1.10e+03
S9810022963	D	Bromide by Ion Chromatograph	ug/mL	98.89	<1.25e-01	<1.28e+03	<1.28e+03	n/a	n/a	94.88	1.28e+03

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002963		Nitrate by IC-Dionex 4000/4500	ug/mL	102.0	<1.39e-01	1.53e-05	1.50e-05	1.51e-05	1.98	101.5	1.42e-03	N/A
S981002963		Phosphate-IC-Dionex 4000/4500	ug/mL	102.7	<1.20e-01	<1.22e-03	<1.22e-03	n/a	n/a	99.63	1.22e-03	N/A
S981002963		Sulfate by IC-Dionex 4000/4500	ug/mL	100.0	<1.38e-01	1.92e-03	2.66e-03	2.29e-03	32.3	100.9	1.41e-03	N/A
S981002963		Oxalate-IC-Dionex 4000/4500	ug/mL	103.3	<1.09e-01	<1.07e-03	<1.07e-03	n/a	n/a	101.1	1.07e-03	N/A
S981002963		Cesium-137 by GEA	uCi/mL	102.1	1.00e-03	2.36e-02	238.0	237.0	0.84	n/a	n/a	0.120
S981002968		Chromium (VI) by Spec.	ug/mL	98.22	<2.58e-01	1.01e-03	986.0	998.0	2.40	95.00	1.00e-02	N/A
S981002968		D I 129 by ICP/MS Acid Addition	ug/mL	106.6	<1.20e-04	< 4.860	<4.86e0	n/a	n/a	n/a	4.860	N/A
S981002968		OH- by Pot. Titration	ug/mL	101.2	n/a	5.10e-04	4.72e-04	4.91e-04	7.74	n/a	5.00e-03	N/A
S981002968		TOC by Persulfate/Coulometry	ug/mL	93.33	13.00	2.62e-03	2.51e-03	2.56e-03	4.29	n/a	40.00	N/A
S981002968		D Uranium-233 by ICP/MS Acid Add	ug/mL	n/a	<3.23e-05	< 1.307	<1.31e0	n/a	n/a	n/a	1.307	N/A
S981002968		D Uranium-234 by ICP/MS Acid Add	ug/mL	n/a	<4.44e-05	< 1.800	<1.80e0	n/a	n/a	n/a	1.800	N/A
S981002968		D Uranium-235 by ICP/MS Acid Add	ug/mL	101.4	<1.64e-05	<6.65e-01	<6.65e-1	n/a	n/a	n/a	6.65e-01	N/A
S981002968		D Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<2.27e-05	<9.21e-01	<9.21e-1	n/a	n/a	n/a	9.21e-01	N/A
S981002968		D Uranium-238 by ICP/MS Acid Add	ug/mL	106.0	<2.14e-05	<8.65e-01	<8.65e-1	n/a	n/a	n/a	8.65e-01	N/A
S981002968		D Tc 99 by ICP/MS Acid Addition	ug/mL	81.22	<2.00e-06	5.464	5.420	5.442	0.74	82.65	8.10e-02	N/A
S981002968		D Sn 126 by ICP/MS Acid Addition	ug/mL	n/a	<1.50e-04	< 6.075	<6.08e0	n/a	n/a	n/a	6.075	N/A
S981002968		A Neptunium-237 by ICP/MS	ug/mL	138.2	n/a	< 6.638	<1.64e0	n/a	n/a	159.5	1.638	N/A
S981002968		A Plutonium-239 by ICP/MS	ug/mL	144.8	<1.85e-05	<7.49e-01	<7.49e-1	n/a	n/a	145.2	7.49e-01	N/A
S981002968		A Plutonium-240 by ICP/MS	ug/mL	n/a	<1.27e-05	<5.14e-01	<5.14e-1	n/a	n/a	n/a	5.14e-01	N/A
S981002968		A Plutonium-242 by ICP/MS	ug/mL	n/a	<7.34e-06	<2.97e-01	<2.97e-1	n/a	n/a	n/a	2.97e-01	N/A
S981002968		A Pu/Am-241 by ICP/MS	ug/mL	n/a	<9.43e-06	<3.83e-01	<3.83e-1	n/a	n/a	n/a	3.83e-01	N/A

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 251
SEGMENT #: 4A

SEGMENT PORTION: L Lower Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002978		Bulk Density of Sample	g/mL	n/a	n/a	DRY	n/a	n/a	n/a	n/a	5.00e-01	n/a
S981002982		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	5.00e-01	n/a
S981002982		DSC Exotherm on Perkin Elmer	Joules/g	99.23	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002982		% Water by TGA on Perkin Elmer	%	99.17	n/a	11.77	10.07	10.92	15.6	n/a	5.000	n/a
S981002982		TIC by Acid/Coulometry	ug/g	101.8	6.100	6.41e+03	5.59e+03	6.00e+03	13.7	n/a	5.000	n/a
S981002982		TIC by Persulfate/Coulometry	ug/g	96.33	15.50	3.72e+02	308.0	340.0	18.8	n/a	40.00	n/a
S981002982		F Strontium-89/90 High Level	uCi/g	98.08	<1.34e-03	4.910	3.500	4.205	33.5	n/a	3.00e+03	9.05E-01
S981002982		F Aluminum - ICP-Fusion	ug/g	97.00	<1.00e-02	3.41e+04	3.08e+04	3.24e+04	10.2	n/a	1.03e+03	n/a
S981002982		F Boron - ICP-Fusion	ug/g	103.2	<5.00e-02	<1.03e+03	<1.03e+03	n/a	n/a	n/a	1.03e+03	n/a
S981002982		F Barium - ICP-Fusion	ug/g	97.00	<5.00e-02	<1.03e+03	<1.03e+03	n/a	n/a	n/a	2.06e+03	n/a
S981002982		F Bismuth - ICP-Fusion	ug/g	100.4	<1.00e-01	3.41e+03	<2.05e+03	2.86e+03	38.9	n/a	2.06e+03	n/a
S981002982		F Calcium - ICP-Fusion	ug/g	100.4	<1.00e-01	3.41e+03	2.30e+03	2.86e+03	38.9	n/a	2.06e+03	n/a
S981002982		F Chromium - ICP-Fusion	ug/g	97.00	<5.00e-02	4.18e+02	277.0	347.5	40.6	n/a	206.0	n/a
S981002982		F Iron - ICP-Fusion	ug/g	97.00	<5.00e-02	3.93e+03	3.35e+03	3.64e+03	15.9	n/a	1.03e+03	n/a
S981002982		F Lithium - ICP-Fusion	ug/g	99.60	<1.00e-02	<2.06e+02	<2.05e+02	n/a	n/a	n/a	206.0	n/a
S981002982		F Manganese - ICP-Fusion	ug/g	99.60	<1.00e-02	<2.06e+02	<2.05e+02	n/a	n/a	n/a	206.0	n/a
S981002982		F Sodium - ICP-Fusion	ug/g	99.00	<1.00e-01	2.13e+05	2.14e+05	2.14e+05	0.47	n/a	2.06e+03	n/a
S981002982		F Nickel - ICP-Fusion	ug/g	100.0	<1.48e-01	4.17e+03	5.75e+03	5.11e+03	25.0	n/a	414.0	n/a
S981002982		F Silicon - ICP-Fusion	ug/g	104.8	<5.00e-02	2.16e+04	1.70e+04	2.03e+04	32.5	n/a	1.03e+04	n/a
S981002982		F Uranium - ICP-Fusion	ug/g	96.90	<5.00e-01	<1.03e+04	<1.03e+04	n/a	n/a	n/a	206.0	n/a
S981002982		F Zinc - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.06e+02	<2.05e+02	n/a	n/a	n/a	206.0	n/a
S981002982		F Zirconium - ICP-Fusion	ug/g	97.60	<1.00e-02	<2.06e+02	<2.05e+02	n/a	n/a	n/a	206.0	n/a
S981002982		F Cesium-137 by GEA	uCi/g	100.3	<1.39e-02	43.78	37.20	40.44	16.6	n/a	n/a	0.430
S981002982		F Alpha of Digested Solid	uCi/g	93.89	<2.65e-01	<1.90e-01	<1.52E-1	n/a	n/a	n/a	3.07e-01	5.00E+02
S981002992		A 1129 by ICP/MS Actidig LA-505159	ug/g	109.0	<1.20e-03	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Uranium-233 by ICP/MS Actidig 159	ug/g	n/a	<1.20e-04	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Uranium-234 by ICP/MS Actidig 159	ug/g	106.0	<1.20e-04	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Uranium-235 by ICP/MS Actidig 159	ug/g	106.0	<1.20e-04	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Uranium-236 by ICP/MS Actidig 159	ug/g	106.0	<1.20e-04	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Uranium-238 by ICP/MS Actidig 159	ug/g	106.0	<1.20e-04	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.71E-01	n/a
S981002992		A Tc99 by ICP/MS Actidig LA-505159	ug/g	86.29	<3.00e-05	5.30e-01	6.10e-01	5.70e-01	14.0	n/a	7.90e-02	n/a
S981002992		A Sr726 by ICP/MS Actidig LA-505159	ug/g	115.4	<1.20e-05	<4.71E-01	<4.78E-01	n/a	n/a	n/a	4.80e-02	n/a
S981002992		A Neptunium-237 by ICP/MS	ug/g	97.30	<3.19e-03	<1.69e-01	<1.51e-01	n/a	n/a	n/a	1.49e-01	n/a
S981002992		A Plutonium-239 by ICP/MS	ug/g	n/a	<9.30e-03	<3.78e-02	<3.78e-02	n/a	n/a	n/a	3.70e-02	n/a
S981002992		A Plutonium-240 by ICP/MS	ug/g	n/a	<4.13e-03	<2.81e-02	<2.85e-02	n/a	n/a	n/a	2.80e-02	n/a
S981002992		A Plutonium-242 by ICP/MS	ug/g	101.4	<1.41e-03	<5.13e-02	<5.13e-02	n/a	n/a	n/a	5.50e-02	n/a
S981002992		A Pu/Am-241 by ICP/MS	ug/g	93.80	<9.00e-02	3.87e-04	3.40e-04	3.64e-04	12.9	n/a	39.10	n/a
S981002992		A Aluminum - ICP-Acid Digest	ug/g	100.4	<4.65e-01	<3.910	<3.9760	n/a	n/a	n/a	39.10	n/a
S981002992		A Boron - ICP-Acid Digest	ug/g	95.20	<5.00e-02	<3.910	<3.9760	n/a	n/a	n/a	39.10	n/a
S981002992		A Barium - ICP-Acid Digest	ug/g	90.60	<1.00e-01	<3.910	<3.9760	n/a	n/a	n/a	78.00	n/a
S981002992		A Bismuth - ICP-Acid Digest	ug/g	102.0	<1.59e-01	1.44e+03	2.68e+03	2.06e+03	40.2	n/a	78.00	n/a
S981002992		A Calcium - ICP-Acid Digest	ug/g	94.00	<1.00e-02	3.56e+02	359.0	357.5	0.84	n/a	7.800	n/a
S981002992		A Chromium - ICP-Acid Digest	ug/g	92.80	<5.10e-02	2.31e+03	2.73e+03	2.52e+03	16.7	n/a	39.10	n/a
S981002992		A Iron - ICP-Acid Digest	ug/g	92.80	<5.10e-02	2.31e+03	2.73e+03	2.52e+03	16.7	n/a	39.10	n/a

Sample#	R	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002992	A	Potassium - ICP-Acid Digest	ug/g	98.60	< 5.00e-01	6.71e+02	691.0	684.0	3.80	n/a	391.0
S981002992	A	Lithium - ICP-Acid Digest	ug/g	93.40	< 1.00e-02	< 7.820	< 9560	n/a	n/a	n/a	7180
S981002992	A	Manganese - ICP-Acid Digest	ug/g	93.40	< 1.00e-02	68.30	84.70	76.0	21.4	n/a	7180
S981002992	A	Sodium - ICP-Acid Digest	ug/g	92.40	< 2.00e-02	1.94e+03	1.97e+03	1.87e+03	1.55	n/a	76.00
S981002992	A	Strontium - ICP-Acid Digest	ug/g	92.40	< 2.00e-02	5.11e+01	5.11e+01	51.50	1.52	n/a	15.40
S981002992	A	Silicon - ICP-Acid Digest	ug/g	150.0	< 7.50e-01	1.23e+03	1.11e+03	1.16e+03	9.72	n/a	391.0
S981002992	A	Zinc - ICP-Acid Digest	ug/g	90.90	< 5.00e-01	3.97e+02	3.97e+02	26.70	n/a	n/a	7180
S981002992	A	Zinc - ICP-Acid Digest	ug/g	85.20	< 2.40e-02	< 7.820	< 9560	26.70	1.72	n/a	7180
S981002992	A	Zirconium - ICP-Acid Digest	ug/g	92.80	< 1.00e-02	1.44e+02	153.0	149.3	4.78	n/a	7180
S981002992	A	Fluoride - ICP-Acid Digest	ug/g	95.80	< 2.00e-02	2.75e+03	3.35e+03	3.05e+03	19.7	n/a	7180
S981002992	A	Chloride - ICP-Acid Digest	ug/g	113.1	< 1.20e-02	6.59e+02	664.0	661.6	0.76	n/a	171.7
S981002992	A	Nitrite - ICP-Acid Digest	ug/g	102.6	< 1.70e-02	7.85e+03	7.63e+03	7.74e+03	2.84	n/a	1.09e+03
S981002992	A	Bromide - ICP-Acid Digest	ug/g	100.3	< 1.20e-01	< 1.20e+03	< 1.20e+03	n/a	n/a	n/a	1.20e+03
S981002992	A	Nitrate by IC-Dioxex 4000/4500	ug/g	97.80	< 1.70e-01	4.45e+05	3.84e+05	4.20e+05	12.2	n/a	1.40e+03
S981002992	A	Phosphate by IC-Dioxex 4000/4500	ug/g	104.9	< 1.20e-01	8.69e+03	1.02e+04	9.44e+03	16.0	n/a	1.21e+03
S981002992	A	Sulfate by IC-Dioxex 4000/4500	ug/g	102.8	< 1.30e-01	5.22e+03	6.04e+03	5.63e+03	14.6	n/a	1.39e+03
S981002992	A	Oxalate - IC-Dioxex 4000/4500	ug/g	103.5	< 1.08e-01	6.46e+03	6.03e+03	6.25e+03	6.89	n/a	852.4

Table 1. Data Summary Report.
BY-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPB %	Spk Rec %	Det Limit	Count Err%
S981002979		Bulk Density of Sample	g/ml	n/a	n/a	n/a	n/a	n/a	n/a	n/a	5.00e-01	n/a
S981002983		DSC Exotherm DRY Calculated	Joules/g DRY	n/a	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002983		DSC Exotherm on Perkin Elmer	Joules/g	99.23	n/a	0.00e+00	0.00e+00	0.00e+00	0.00	n/a	n/a	n/a
S981002983		% Water by TGA on Perkin Elmer	%	99.17	n/a	36.45	43.77	40.11	18.2	n/a	n/a	n/a
S981002983		TIC by Acid/Coulometry	ug/g	96.68	4.60	4.09e+03	4.25e+03	4.17e+03	3.84	n/a	5.000	n/a
S981002983		TOC by Acid/Coulometry	ug/g	93.67	2.100	1.50e+03	1.91e+03	1.70e+03	24.0	n/a	40.00	n/a
S981002983		F Strontium-89/90 High Level	uCi/g	99.04	7.00e-03	3.870	3.760	3.815	2.88	n/a	5.00e-03	1.47e+00
S981002989		F Aluminum-10P-Fusion	ug/g	97.00	<5.00e-02	2.47e+04	2.61e+04	2.54e+04	5.51	n/a	1.02e+03	n/a
S981002989		F Boron -10P-Fusion	ug/g	103.2	<5.00e-02	4.02e+03	<1.01e3	n/a	n/a	n/a	1.02e+03	n/a
S981002989		F Barium -10P-Fusion	ug/g	97.00	<5.00e-02	4.02e+03	<1.01e3	n/a	n/a	n/a	1.02e+03	n/a
S981002989		F Bismuth -10P-Fusion	ug/g	100.6	<1.00e-01	<2.03e+03	<2.03e+03	n/a	n/a	n/a	2.03e+03	n/a
S981002989		F Calcium -10P-Fusion	ug/g	104.4	<1.00e-01	2.92e+03	2.34e+03	2.63e+03	22.1	n/a	2.03e+03	n/a
S981002989		F Chromium -10P-Fusion	ug/g	100.4	<1.00e-01	5.21e+02	481.0	501.0	7.98	n/a	203.0	n/a
S981002989		F Iron -10P-Fusion	ug/g	97.00	<1.00e-02	1.59e+03	1.91e+03	1.75e+03	18.3	n/a	1.02e+03	n/a
S981002989		F Lithium -10P-Fusion	ug/g	99.60	<1.00e-02	<2.03e+02	<2.03e2	n/a	n/a	n/a	203.0	n/a
S981002989		F Manganese -10P-Fusion	ug/g	99.40	<1.00e-01	2.09e+05	2.14e+05	2.12e+05	2.36	n/a	203.0	n/a
S981002989		F Nickel -10P-Fusion	ug/g	99.00	<1.00e-01	5.34e+03	3.81e+03	4.58e+03	33.4	n/a	2.03e+03	n/a
S981002989		F Sodium -10P-Fusion	ug/g	100.8	<5.00e-02	1.01e+04	1.15e+04	1.08e+04	13.0	n/a	406.0	n/a
S981002989		F Silicon -10P-Fusion	ug/g	104.4	<1.00e-01	<1.02e+04	<1.01e4	n/a	n/a	n/a	1.02e+03	n/a
S981002989		F Uranium -10P-Fusion	ug/g	96.90	<1.00e-02	<2.03e+02	<2.03e2	n/a	n/a	n/a	203.0	n/a
S981002989		F Zinc -10P-Fusion	ug/g	97.60	<1.00e-02	<2.03e+02	<2.03e2	n/a	n/a	n/a	203.0	n/a
S981002989		F Zirconium -10P-Fusion	ug/g	100.3	<1.39e-02	55.47	56.40	55.94	1.61	n/a	0.360	n/a
S981002989		F Cesium-137 by GEA	uCi/g	93.89	<2.2e-01	<1.87e-01	<2.22e-1	n/a	n/a	n/a	3.02e-01	5.00e+02
S981002989		F Alpha of Digested Solid	uCi/g	109.0	<1.20e-03	<4.811	<4.950	n/a	n/a	n/a	4.810	n/a
S981002993		A 1129 by TCP/MS Actidig LA505159	ug/g	107.0	<1.20e-03	<4.81e-01	<4.95e-1	n/a	n/a	n/a	4.81e-01	n/a
S981002993		A Uranium-233 by TCP/MS Actidig159	ug/g	n/a	<1.20e-04	<4.81e-01	<4.95e-1	n/a	n/a	n/a	4.81e-01	n/a
S981002993		A Uranium-234 by TCP/MS Actidig159	ug/g	106.0	<1.20e-04	8.74e-01	9.35e-01	9.35e-01	13.2	113.9	4.81e-01	n/a
S981002993		A Uranium-235 by TCP/MS Actidig159	ug/g	n/a	<1.60e-04	<6.41e-01	<6.60e-1	n/a	n/a	n/a	6.41e-01	n/a
S981002993		A Uranium-236 by TCP/MS Actidig159	ug/g	106.0	<1.20e-04	1.18e+02	123.0	120.5	4.15	115.0	4.81e-01	n/a
S981002993		A Uranium-238 by TCP/MS Actidig159	ug/g	86.29	<3.30e-05	1.010	1.140	1.075	12.1	n/a	8.00e-02	n/a
S981002993		A Tc99 by TCP/MS Actidig LA505159	ug/g	n/a	<1.50e-03	<6.013	<6.190	n/a	n/a	n/a	6.014	n/a
S981002993		A Sm126 by TCP/MS Actidig LA-505159	ug/g	117.0	<1.21e-03	<4.87e-02	<5.01e-2	n/a	n/a	113.4	4.90e-02	n/a
S981002993		A Neptunium-237 by TCP/MS	ug/g	97.30	<1.79e-03	<1.52e-01	<1.56e-1	n/a	n/a	97.1	1.52e-01	n/a
S981002993		A Plutonium-239 by TCP/MS	ug/g	n/a	<9.50e-04	<3.81e-02	<3.92e-2	n/a	n/a	n/a	3.80e-02	n/a
S981002993		A Plutonium-240 by TCP/MS	ug/g	n/a	<1.15e-04	<2.84e-02	<2.95e-2	n/a	n/a	n/a	2.90e-02	n/a
S981002993		A Plutonium-242 by TCP/MS	ug/g	n/a	<1.41e-05	<5.65e-02	<5.81e-2	n/a	n/a	102.8	5.60e-02	n/a
S981002993		A Pu/Am-241 by TCP/MS	ug/g	93.60	9.41e-02	1.92e+04	1.95e+04	1.88e+04	6.90	n/a	39.90	n/a
S981002993		A Aluminum -10P-Acid Digest	ug/g	100.4	<1.83e-02	51.40	81.80	66.60	45.6	89.80	39.90	n/a
S981002993		A Boron -10P-Acid Digest	ug/g	95.20	<3.00e-02	<39.90	<4.11e1	n/a	n/a	88.80	39.90	n/a
S981002993		A Barium -10P-Acid Digest	ug/g	102.0	<1.00e-01	<79.80	<8.21e1	n/a	n/a	88.80	79.80	n/a
S981002993		A Bismuth -10P-Acid Digest	ug/g	106.0	<1.59e-01	1.01e+03	2.35e+03	1.68e+03	79.8	113.2	79.60	n/a
S981002993		A Calcium -10P-Acid Digest	ug/g	92.80	<1.00e-02	4.23e+02	4.87e+02	454.5	14.3	90.80	7.960	n/a
S981002993		A Chromium -10P-Acid Digest	ug/g	92.80	<1.00e-02	1.22e+03	988.0	1.01e+03	4.16	122.2	39.90	n/a
S981002993		A Iron -10P-Acid Digest	ug/g	92.80	<1.00e-02	1.22e+03	988.0	1.01e+03	4.16	122.2	39.90	n/a

150

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T02993	A	Potassium - ICP-Acid Digest	ug/g	86.60	< 5.00e-01	9.50e-02	929.0	959.5	6.36	75.60	399.0	n/a
S98T02993	A	Lithium - ICP-Acid Digest	ug/g	93.40	< 1.00e-02	< 7.980	< 8.21e0	< 7.980	n/a	85.00	7.960	n/a
S98T02993	A	Manganese - ICP-Acid Digest	ug/g	93.00	< 1.00e-02	27.70	28.20	27.95	1.79	87.80	7.960	n/a
S98T02993	A	Sodium - ICP-Acid Digest	ug/g	117.4	7.43e-01	1.92e+05	2.08e+05	2.00e+05	8.00	-1.790e3	79.60	n/a
S98T02993	A	Nickel - ICP-Acid Digest	ug/g	92.40	< 2.00e-02	59.80	62.60	61.20	4.58	89.80	16.00	n/a
S98T02993	A	Silicon - ICP-Acid Digest	ug/g	150.0	7.58e-01	9.81e-02	1.13e+03	1.06e+03	14.1	118.2	39.90	n/a
S98T02993	A	Uranium - ICP-Acid Digest	ug/g	90.90	< 5.00e-01	< 3.95e-02	< 4.11e2	n/a	n/a	89.90	399.0	n/a
S98T02993	A	Zinc - ICP-Acid Digest	ug/g	86.20	3.60e-02	< 7.980	< 8.21e0	n/a	n/a	90.60	7.960	n/a
S98T02993	A	Zirconium - ICP-Acid Digest	ug/g	92.80	< 1.00e-02	< 7.980	< 8.21e0	n/a	n/a	87.20	7.960	n/a
S98T02997	W	Formate by IC-Dionex 4000/4500	ug/g	95.80	< 2.00e-02	2.13e-02	199.0	205.8	6.80	n/a	73.20	n/a
S98T02997	W	Fluoride by IC-Dionex 4000/4500	ug/g	102.4	< 1.20e-02	1.62e+03	1.60e+03	1.61e+03	1.24	n/a	120.3	n/a
S98T02997	W	Chloride by IC-Dionex 4000/4500	ug/g	102.4	< 1.70e-02	1.20e+03	985.0	1.09e+03	19.7	n/a	170.6	n/a
S98T02997	W	Nitrate by IC-Dionex 4000/4500	ug/g	102.4	5.22e-01	1.32e+04	1.22e+04	1.27e+04	7.87	n/a	1.08e+03	n/a
S98T02997	W	Bromide by Ion Chromatograph	ug/g	106.6	1.25e-01	4.95e+03	5.26e+03	5.11e+03	6.07	n/a	2.16e+03	n/a
S98T02997	W	Nitrate by IC-Dionex 4000/4500	ug/g	104.6	1.25e-01	4.95e+03	5.26e+03	5.11e+03	6.07	n/a	2.16e+03	n/a
S98T02997	W	Phosphate by IC-Dionex 4000/4500	ug/g	104.6	< 1.20e-01	5.71e+03	5.97e+03	5.84e+03	16.9	n/a	1.26e+03	n/a
S98T02997	W	Sulfate by IC-Dionex 4000/4500	ug/g	102.9	< 1.20e-01	5.00e+03	5.37e+03	5.19e+03	16.9	n/a	1.35e+03	n/a
S98T02997	W	Oxalate by IC-Dionex 4000/4500	ug/g	106.3	< 1.00e-01	2.62e+03	2.30e+03	2.56e+03	4.69	n/a	1.09e+03	n/a

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T02959	DSC	Exochrom DTY, Calculated	Joules/g DTY	n/a	n/a	23.97	23.44	23.70	2.24	n/a	n/a	n/a
S98T02959	DSC	Exochrom on Perkin Elmer	Joules/g	99.19	n/a	13.19	12.30	13.04	2.22	n/a	n/a	n/a
S98T02959	Ammonia	by ISE-Std Additions	ug/mL	98.12	< 10.00	< 1.00e+1	n/a	n/a	n/a	n/a	10.00	n/a
S98T02959	Specific Gravity		Sp. G.	98.99	n/a	1.455	1.435	1.445	1.38	n/a	1.00e-03	n/a
S98T02959	% Water by Tit on Perkin Elmer		%	98.86	n/a	45.28	44.66	44.97	1.38	n/a	30.10	n/a
S98T02959	Alpha In Liquid Samples		uCi/mL	94.44	< 5.75e-03	1.83e-02	2.34e-02	2.09e-02	24.5	n/a	2.20e-02	8.78e+01
S98T02964	Strontium-89/90 High Level		uCi/mL	110.7	2.77e-04	8.00e-02	8.12e-02	8.06e-02	1.49	n/a	2.20e-04	2.31e+00
S98T02964	Pu-239/240 by TRU-SPEC Resin		uCi/mL	97.80	< 8.17e-06	8.13e-06	< 8.01E-6	n/a	n/a	n/a	8.20e-06	7.75e+00
S98T02964	Aluminum-ICP-Acid Dil.		ug/mL	97.80	< 5.00e-02	4.58e+04	4.61e+04	4.60e+04	0.65	n/a	30.10	n/a
S98T02964	Boron-ICP-Acid Dil.		ug/mL	102.4	< 5.00e-02	< 30.10	< 3.01e1	n/a	n/a	n/a	30.10	n/a
S98T02964	Barium-ICP-Acid Dil.		ug/mL	97.00	< 5.00e-02	< 30.10	< 3.01e1	n/a	n/a	n/a	30.10	n/a
S98T02964	Bismuth-ICP-Acid Dil.		ug/mL	100.4	< 1.00e-01	< 60.10	< 6.01e1	n/a	n/a	n/a	60.10	n/a
S98T02964	Calcium-ICP-Acid Dil.		ug/mL	103.2	< 1.00e-01	< 60.10	< 6.01e1	n/a	n/a	n/a	60.10	n/a
S98T02964	Chromium-ICP-Acid Dil.		ug/mL	99.80	< 1.00e-02	1.31e+03	1.32e+03	1.32e+03	0.76	n/a	6.010	n/a
S98T02964	Iron-ICP-Acid Dil.		ug/mL	96.40	< 5.00e-02	39.00	40.40	39.70	3.53	n/a	30.10	n/a
S98T02964	Potassium-ICP-Acid Dil.		ug/mL	94.40	< 5.00e-02	4.77e+03	4.76e+03	4.76e+03	0.21	n/a	300.0	n/a
S98T02964	Lithium-ICP-Acid Dil.		ug/mL	99.80	< 1.00e-02	< 6.010	< 6.01e0	n/a	n/a	n/a	6.010	n/a
S98T02964	Manganese-ICP-Acid Dil.		ug/mL	98.60	< 1.00e-02	< 6.010	< 6.01e0	n/a	n/a	n/a	6.010	n/a
S98T02964	Sodium-ICP-Acid Dil.		ug/mL	98.80	< 1.00e-01	2.09e+05	2.10e+05	2.10e+05	0.48	n/a	60.10	n/a
S98T02964	Nickel-ICP-Acid Dil.		ug/mL	98.80	< 2.00e-02	< 12.00	< 1.20e1	n/a	n/a	n/a	12.00	n/a
S98T02964	Silicon-ICP-Acid Dil.		ug/mL	104.0	< 5.00e-02	2.07e+02	2.15e+02	211.0	3.79	n/a	300.0	n/a
S98T02964	Uranium-ICP-Acid Dil.		ug/mL	96.50	< 5.00e-02	< 3.01e02	< 3.01e2	n/a	n/a	n/a	6.010	n/a
S98T02964	Zinc-ICP-Acid Dil.		ug/mL	97.20	< 1.00e-02	< 6.010	< 6.01e0	n/a	n/a	n/a	6.010	n/a
S98T02964	Zirconium-ICP-Acid Dil.		ug/mL	94.80	< 2.00e-02	1.77e+03	1.72e+03	1.74e+03	2.87	n/a	37.57	n/a
S98T02964	Formate by Ion Chromatograph		ug/mL	107.6	< 2.00e-02	< 49.69	< 4.97e1	n/a	n/a	n/a	49.69	n/a
S98T02964	Fluoride by IC-Dionex 4000/4500		ug/mL	95.00	< 1.00e-02	5.22e+03	5.33e+03	5.28e+03	2.09	n/a	70.40	n/a
S98T02964	Chloride by IC-Dionex 4000/4500		ug/mL	103.9	< 1.70e-02	7.76e+04	7.78e+04	7.77e+04	1.55	n/a	447.2	n/a
S98T02964	Nitrite by Ion Chromatograph		ug/mL	99.46	< 1.25e-01	< 5.18e+02	< 5.18e2	n/a	n/a	n/a	517.6	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002964		Nitrate by IC-Dionex 4000/4500	ug/mL	102.6	<1.35e-01	1.74e-03	1.74e-03	1.74e-03	3.0	N/A	5.7e-6	N/A
S981002964		Phosphate-IC-Dionex 4000/4500	ug/mL	102.3	<1.20e-01	7.23e-02	7.29e-03	1.79e-03	3.2	N/A	4.9e-9	N/A
S981002964		Sulfate by IC-Dionex 4000/4500	ug/mL	79.14	<1.38e-01	4.69e-03	1.83e-03	1.76e-03	7.93	N/A	5.71e-5	N/A
S981002964		Oxalate-IC-Dionex 4000/4500	ug/mL	102.4	<1.05e-01	2.52e-02	4.85e-03	4.85e-03	N/A	N/A	4.3e-3	N/A
S981002964		Cesium-137 by GEA	LCI/mL	102.4	<1.05e-01	2.52e-02	2.22e-03	2.22e-03	0.00	N/A	N/A	0.120
S981002969		Chromium (VI) by Spec.	ug/mL	98.22	<2.58e-03	1.24e-03	1.24e-03	1.25e-03	1.40	N/A	1.00e-02	N/A
S981002969		On- By Pot. Titration	ug/mL	103.2	<1.20e-04	4.78e-04	4.78e-04	5.26e-04	18.8	N/A	4.86e-6	N/A
S981002969		TOC by Persulfate/Conductivity	ug/mL	103.2	<1.20e-04	4.78e-04	2.95e-03	2.97e-03	0.67	N/A	5.00e-03	N/A
S981002969		Uranium-235 by ICP/MS Acid Add	ug/mL	88.33	<3.23e-05	2.98e-02	<1.31e-03	N/A	N/A	N/A	1.307	N/A
S981002969		Uranium-234 by ICP/MS Acid Add	ug/mL	N/A	<4.44e-05	<1.80e-01	<6.45e-01	N/A	N/A	N/A	1.800	N/A
S981002969		Uranium-233 by ICP/MS Acid Add	ug/mL	101.4	<1.44e-05	<6.63e-01	<9.21e-01	N/A	N/A	N/A	6.65e-01	N/A
S981002969		Uranium-235 by ICP/MS Acid Add	ug/mL	101.4	<2.27e-05	<9.21e-01	<9.21e-01	N/A	N/A	N/A	9.21e-01	N/A
S981002969		Uranium-238 by ICP/MS Acid Add	ug/mL	106.0	<2.16e-05	<8.65e-01	<8.65e-01	N/A	N/A	N/A	8.65e-01	N/A
S981002969		Uranium-238 by ICP/MS Acid Add	ug/mL	81.22	<2.00e-04	<5.512	5.450	5.481	1.09	N/A	8.10e-02	N/A
S981002969		U-238 by ICP/MS Acid Addition	ug/mL	N/A	<1.50e-04	<6.075	<6.08e-04	N/A	N/A	N/A	6.075	N/A
S981002969		Strontium-90 by ICP/MS	ug/mL	138.2	<1.50e-04	<6.075	<6.08e-04	N/A	N/A	N/A	6.075	N/A
S981002969		Strontium-90 by ICP/MS	ug/mL	144.8	<1.85e-05	<7.49e-01	<7.49e-01	N/A	N/A	N/A	7.49e-01	N/A
S981002969		Plutonium-239 by ICP/MS	ug/mL	144.8	<1.27e-05	<5.14e-01	<5.14e-01	N/A	N/A	N/A	5.14e-01	N/A
S981002969		Plutonium-240 by ICP/MS	ug/mL	N/A	<7.34e-06	<2.97e-01	<2.97e-01	N/A	N/A	N/A	2.97e-01	N/A
S981002969		Plutonium-242 by ICP/MS	ug/mL	N/A	<7.34e-06	<2.97e-01	<2.97e-01	N/A	N/A	N/A	2.97e-01	N/A
S981002969		Pu/Am-241 by ICP/MS	ug/mL	N/A	<9.45e-06	<3.83e-01	<3.83e-01	N/A	N/A	N/A	3.83e-01	N/A

CORE NUMBER: 252
SEGMENT #: 5

Table 1. Data Summary Report.
8Y-105 (2)

SEGMENT PORTION: Drinable Liquid

Sample #	#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003002		USC Exotherm Dry Calculated	Joules/g Dry	n/a	n/a	44.96	66.98	55.97	39.3	n/a	n/a	n/a
S98T003002		USC Exotherm on Rockin Elmer	Joules/g	100.6	n/a	21.37	31.84	26.61	39.4	n/a	n/a	n/a
S98T003002		Ammonia by ISE-Std Additions	ug/mL	117.3	<5.00e-02	6.49e-02	580.0	614.5	11.2	n/a	55.00-03	n/a
S98T003002		Specific Gravity	Sp.G.	98.99	n/a	1.436	1.437	1.437	0.07	n/a	1.00e-03	n/a
S98T003002		% Water by Grav on Rockin Elmer	ug/mL	98.99	n/a	53.24	51.69	52.47	2.95	n/a	1.30e-02	3.59e+01
S98T003002		Alpha in Liquid Samples	ug/mL	93.89	<9.21e-03	4.92e-02	4.40e-02	4.69e-02	9.81	n/a	1.30e-02	3.59e+01
S98T003002		Strontium-90/90m by Level	ug/mL	101.9	<2.27e-04	1.17e-01	1.21e-01	1.19e-01	3.36	n/a	2.56e-04	1.94e+00
S98T003002		Pu-239/240 by ThU-SPEC Resin	ug/mL	106.2	<2.99e-05	3.60e-05	3.57E-5	n/a	n/a	n/a	3.60e-05	1.27e+01
S98T003006		Aluminum-CP-Acid DIL	ug/mL	95.00	<5.00e-02	3.74e+04	3.74e+04	3.77e+04	1.59	n/a	30.10	n/a
S98T003006		Boron-CP-Acid DIL	ug/mL	101.2	<5.00e-02	<30.10	<3.01e1	n/a	n/a	n/a	30.10	n/a
S98T003006		Barium-CP-Acid DIL	ug/mL	94.60	<5.00e-02	<30.10	<6.01e1	n/a	n/a	n/a	30.10	n/a
S98T003006		Bismuth-CP-Acid DIL	ug/mL	99.60	<1.00e-01	<60.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T003006		Calcium-CP-Acid DIL	ug/mL	103.4	<1.00e-01	<60.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T003006		Chromium-CP-Acid DIL	ug/mL	99.60	<5.00e-02	1.29e+03	1.28e+03	1.28e+03	1.56	n/a	30.10	n/a
S98T003006		Iron-CP-Acid DIL	ug/mL	95.40	<5.00e-02	33.10	32.70	32.90	1.22	n/a	300.0	n/a
S98T003006		Potassium-CP-Acid DIL	ug/mL	93.20	<5.00e-01	4.25e+03	4.15e+03	4.20e+03	2.58	n/a	6.010	n/a
S98T003006		Lithium-CP-Acid DIL	ug/mL	96.00	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003006		Manganese-CP-Acid DIL	ug/mL	97.80	<1.00e-01	2.09e+05	2.04e+05	2.07e+05	2.42	n/a	12.00	n/a
S98T003006		Sodium-CP-Acid DIL	ug/mL	99.40	<5.00e-02	<12.00	<1.20e1	n/a	n/a	n/a	12.00	n/a
S98T003006		Nickel-CP-Acid DIL	ug/mL	103.2	<5.00e-02	1.62e+02	155.0	158.5	4.42	n/a	300.0	n/a
S98T003006		Silicon-CP-Acid DIL	ug/mL	94.60	<5.00e-01	<6.010	<3.01e2	n/a	n/a	n/a	6.010	n/a
S98T003006		Uranium-CP-Acid DIL	ug/mL	95.60	<5.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003006		Zinc-CP-Acid DIL	ug/mL	96.20	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003006		Formate by Ion Chromatograph	ug/mL	101.7	<1.00e-02	2.28e+02	1.30e+03	1.29e+03	1.55	n/a	131.5	n/a
S98T003006		Fluoride-IC-Dionex 4000/4500	ug/mL	91.41	<1.20e-02	2.98e+02	280.0	288.8	6.23	n/a	87.57	n/a
S98T003006		Chloride-IC-Dionex 4000/4500	ug/mL	102.2	<1.70e-02	5.12e+03	5.00e+03	5.06e+03	2.37	n/a	556.3	n/a
S98T003006		Nitrite-IC-Dionex 4000/4500	ug/mL	101.4	<1.08e-01	6.79e+04	6.71e+04	6.84e+04	3.66	n/a	643.9	n/a
S98T003006		Bromide by Ion Chromatograph	ug/mL	99.68	<1.25e-01	7.14e+02	704.0	708.8	1.41	n/a	714.0	n/a
S98T003006		Nitrate by Ion Chromatograph	ug/mL	100.6	<1.20e-01	<6.18e+02	<6.18e2	n/a	n/a	n/a	618.0	n/a
S98T003006		Phosphate by IC-Dionex 4000/4500	ug/mL	101.6	<1.38e-01	2.00e+03	2.02e+03	2.01e+03	1.00	n/a	710.8	n/a
S98T003006		Sulfate by IC-Dionex 4000/4500	ug/mL	104.8	<1.05e-01	<5.41e+02	<5.41e2	n/a	n/a	n/a	540.9	n/a
S98T003006		Oxalate-IC-Dionex 4000/450	ug/mL	100.8	<2.56e-01	2.10e+02	210.0	210.0	0.00	n/a	1.00e-02	n/a
S98T003010		Chromium (VI) by Spec	ug/mL	105.0	<2.56e-01	1.11e+03	1.11e+03	1.11e+03	0.00	n/a	1.00e-02	n/a
S98T003010		1129 by ICP/MS Acid Addition	ug/mL	98.84	<2500.0	4.86e0	3.91e+04	4.04e+04	6.20	n/a	2.50e+03	n/a
S98T003010		ThC by Pot. Titration	ug/mL	94.33	<1.700	2.35e+03	2.29e+03	2.34e+03	3.85	n/a	4.100	n/a
S98T003010		ThC by Persulfate/Oxometry	ug/mL	n/a	<3.23e-05	<1.307	<1.31e0	n/a	n/a	n/a	1.307	n/a
S98T003010		Uranium-235 by ICP/MS Acid Add	ug/mL	101.4	<4.44e-05	<1.800	<6.65e-1	n/a	n/a	n/a	6.65e-01	n/a
S98T003010		Uranium-236 by ICP/MS Acid Add	ug/mL	n/a	<2.27e-05	<9.21e-1	<8.65e-1	n/a	n/a	n/a	9.21e-01	n/a
S98T003010		Uranium-238 by ICP/MS Acid Add	ug/mL	106.0	<2.14e-05	<8.65e-1	<8.65e-1	n/a	n/a	n/a	8.65e-01	n/a
S98T003010		Uranium-238 by ICP/MS Acid Add	ug/mL	81.22	<2.00e-06	4.832	4.830	4.831	0.00	n/a	8.10e-02	n/a

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003010	D	Sn 126	by ICP/MS Acid Addition	ug/mL		< 1.50e-04	< 6.07e-04	< 6.07e-04	n/a	n/a	n/a	6.07e-04	n/a
S98T003010	A	Neptunium-237	by ICP/MS	ug/mL	138.2	n/a	< 1.63e-01	< 1.64e-01	n/a	n/a	n/a	1.63e-01	n/a
S98T003010	A	Plutonium-239	by ICP/MS	ug/mL	144.8	< 1.85e-03	< 7.49e-01	< 7.49e-01	n/a	n/a	n/a	7.49e-01	n/a
S98T003010	A	Plutonium-240	by ICP/MS	ug/mL	n/a	< 1.27e-03	< 5.14e-01	< 5.14e-01	n/a	n/a	n/a	5.14e-01	n/a
S98T003010	A	Plutonium-242	by ICP/MS	ug/mL	n/a	< 7.34e-06	< 2.97e-01	< 2.97e-01	n/a	n/a	n/a	2.97e-01	n/a
S98T003010	A	Pu/Am-241	by ICP/MS	ug/mL	n/a	< 9.43e-06	< 3.83e-01	< 3.83e-01	n/a	n/a	n/a	3.83e-01	n/a

Table 1. Data Summary Report.
BY-105 (2)CORE NUMBER: 252
SEGMENT #: 6

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	Spk Rec	Det Limit	Count Exp
S98T03003	DSC	Exotherm DRY, Calculated	Joules/g DRY	N/A	N/A	1.09e+02	125.1	117.2	13.4	N/A	N/A	N/A
S98T03003	DSC	Exotherm on Perkin Elmer	Joules/g	100.6	N/A	53.02	60.64	56.83	13.4	N/A	N/A	N/A
S98T03003	Ammonia	by ISE-std Additions	ug/mL	117.3	< 5.00	< 55.00	55.40	56.83	13.4	N/A	55.00	N/A
S98T03003	Specific Gravity		Sp.G.	98.99	N/A	1.447	1.438	1.442	0.62	N/A	1.00e-03	N/A
S98T03003	% Water by IGA on Perkin Elmer		%	98.99	N/A	51.64	51.42	51.53	0.43	N/A	1.00e-03	N/A
S98T03003	Alpha In Liquid Samples		uCi/mL	93.89	< 9.21e-03	4.58e-02	4.11e-02	4.35e-02	10.8	N/A	1.30e-02	3.42e+01
S98T03007	Pu-239/240 by TRU-SPEC Resin		uCi/mL	106.2	< 2.99e-05	1.13e-01	1.18e-01	1.15e-01	4.33	N/A	2.49e-04	1.97e+00
S98T03007	D Boron-10/CP-Acid DIL.		ug/mL	95.00	< 5.00e-02	3.80e+04	3.74e+04	3.77e+04	1.59	N/A	3.01e-05	1.00e-02
S98T03007	D Boron-10/CP-Acid DIL.		ug/mL	101.2	< 5.00e-02	< 30.10	< 3.01e1	N/A	N/A	N/A	30.10	N/A
S98T03007	D Barium-137/CP-Acid DIL.		ug/mL	94.60	< 5.00e-02	< 60.10	< 6.01e1	N/A	N/A	N/A	60.10	N/A
S98T03007	D Bismuth-210/CP-Acid DIL.		ug/mL	99.60	< 1.00e-01	< 60.10	< 6.01e1	N/A	N/A	N/A	60.10	N/A
S98T03007	D Calcium-45/CP-Acid DIL.		ug/mL	103.4	< 1.00e-01	< 60.10	< 6.01e1	N/A	N/A	N/A	60.10	N/A
S98T03007	D Chromium-51/CP-Acid DIL.		ug/mL	99.60	< 1.00e-01	1.26e+03	1.24e+03	1.25e+03	1.60	N/A	6.010	N/A
S98T03007	D Iron-59/CP-Acid DIL.		ug/mL	95.40	< 5.00e-02	32.80	31.90	32.35	2.78	N/A	30.10	N/A
S98T03007	D Lithium-7/CP-Acid DIL.		ug/mL	96.00	< 1.00e-01	< 6.010	< 6.01e0	4.28e+03	2.57	N/A	300.0	N/A
S98T03007	D Potassium-40/CP-Acid DIL.		ug/mL	93.20	< 1.00e-02	< 6.010	< 6.01e0	N/A	N/A	N/A	6.010	N/A
S98T03007	D Manganese-54/CP-Acid DIL.		ug/mL	97.80	< 1.00e-01	2.06e+05	2.05e+05	2.06e+05	0.49	N/A	60.10	N/A
S98T03007	D Sodium-24/CP-Acid DIL.		ug/mL	99.40	< 2.00e-02	< 12.00	< 1.20e1	N/A	N/A	N/A	12.00	N/A
S98T03007	D Nickel-63/CP-Acid DIL.		ug/mL	103.2	< 5.00e-02	1.31e+02	1.26e2	128.5	3.89	N/A	30.10	N/A
S98T03007	D Silicon-31/CP-Acid DIL.		ug/mL	94.60	< 1.00e-02	< 3.01e2	< 3.01e2	N/A	N/A	N/A	300.0	N/A
S98T03007	D Uranium-235/CP-Acid DIL.		ug/mL	95.60	< 1.00e-02	< 6.010	< 6.01e0	N/A	N/A	N/A	6.010	N/A
S98T03007	D Zinc-65/CP-Acid DIL.		ug/mL	96.20	< 1.00e-02	< 6.010	< 6.01e0	N/A	N/A	N/A	6.010	N/A
S98T03007	D Zirconium-95/CP-Acid DIL.		ug/mL	90.76	< 1.00e-02	1.13e+03	1.14e+03	1.14e+03	0.88	N/A	131.5	N/A
S98T03007	Formate by Ion Chromatograph		ug/mL	91.41	< 1.20e-02	2.98e+02	291.0	294.3	2.38	N/A	61.81	N/A
S98T03007	Fluoride-IC-Bioxon 4000/4500		ug/mL	102.2	< 1.20e-02	4.64e+03	4.68e+03	4.66e+03	0.86	N/A	87.57	N/A
S98T03007	Chloride-IC-Bioxon 4000/4500		ug/mL	101.4	< 1.08e-01	6.52e+04	6.55e+04	6.53e+04	0.46	N/A	556.3	N/A
S98T03007	Nitrate by Ion Chromatograph		ug/mL	99.68	< 1.25e-01	7.16e+02	664.0	690.0	7.54	N/A	643.9	N/A
S98T03007	Nitrate by IC-Bioxon 4000/4500		ug/mL	101.4	< 1.39e-01	2.01e+05	2.01e+05	2.01e+05	0.00	N/A	716.0	N/A
S98T03007	Phosphate by IC-Bioxon 4000/4500		ug/mL	100.6	< 1.20e-01	1.66e+03	1.68e3	2.15e+03	1.86	N/A	618.1	N/A
S98T03007	Sulfate by IC-Bioxon 4000/4500		ug/mL	104.8	< 1.38e-01	2.13e+03	2.17e+03	2.15e+03	1.86	N/A	710.8	N/A
S98T03007	Oxalate-IC-Bioxon 4000/4500		ug/mL	101.8	< 1.05e-01	< 5.41e2	< 5.41e2	N/A	N/A	N/A	540.9	N/A
S98T03007	Cesium-137 by GEA		uCi/mL	100.4	< 2.50e-03	1.98e+02	200.0	199.0	1.01	N/A	N/A	0.0900
S98T03007	Chromium (VI) by Spec.		ug/mL	100.4	< 2.50e-03	1.10e+03	1.11e+03	1.10e+03	0.90	N/A	1.00e-02	N/A
S98T03007	Oh- by Pot. Titration		ug/mL	105.0	< 1.00e-04	4.860	4.86e0	N/A	N/A	N/A	4.860	N/A
S98T03007	TOC by Persulfate/Coulometry		ug/mL	98.84	< 25.00	4.33e+04	4.47e+04	4.40e+04	3.18	N/A	2.50e+03	N/A
S98T03007	D Uranium-233 by IC/MS Acid Add		ug/mL	94.33	< 1.700	2.33e+03	1.12e+03	1.72e+03	70.1	N/A	40.00	N/A
S98T03007	D Uranium-235 by IC/MS Acid Add		ug/mL	n/a	< 5.19e-05	< 2.101	< 2.10e0	N/A	N/A	N/A	2.101	N/A
S98T03007	D Uranium-235 by IC/MS Acid Add		ug/mL	n/a	< 2.79e-05	< 1.128	< 1.13e0	N/A	N/A	N/A	1.128	N/A
S98T03007	D Uranium-235 by IC/MS Acid Add		ug/mL	102.8	< 2.79e-05	< 1.132	< 1.13e0	N/A	N/A	N/A	1.132	N/A
S98T03007	D Uranium-236 by IC/MS Acid Add		ug/mL	n/a	< 1.271	< 1.271	< 1.27e0	N/A	N/A	N/A	1.271	N/A
S98T03007	D Uranium-236 by IC/MS Acid Add		ug/mL	102.0	< 1.90e-05	< 7.68e-01	< 7.68e-01	N/A	N/A	N/A	7.68e-01	N/A
S98T03007	D Tc 99 by IC/MS Acid Addition		ug/mL	81.73	< 2.00e-06	4.753	4.910	4.832	3.31	N/A	8.10e-02	N/A

Sample#	P /#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T03011	D	Sn 126 by ICP/MS	Acid Addition									
S98T03011	A	NaCl 237 by ICP/MS	ug/mL	138.2	< 1.50e-04	< 6.075	< 6.08e0	D/A	D/A	D/A	6.075	D/A
S98T03011	A	NaCl 237 by ICP/MS	ug/mL	144.8	< 1.85e-03	< 1.638	< 1.64e0	D/A	D/A	D/A	1.638	D/A
S98T03011	A	NaCl 239 by ICP/MS	ug/mL	144.8	< 1.85e-03	< 7.49e-01	< 7.49e-01	D/A	D/A	D/A	7.49e-01	D/A
S98T03011	A	NaCl 240 by ICP/MS	ug/mL	144.8	< 1.27e-03	< 5.14e-01	< 5.14e-01	D/A	D/A	D/A	5.14e-01	D/A
S98T03011	A	NaCl 242 by ICP/MS	ug/mL	144.8	< 7.36e-02	< 2.97e-01	< 2.97e-01	D/A	D/A	D/A	2.97e-01	D/A
S98T03011	A	NaCl 241 by ICP/MS	ug/mL	144.8	< 9.45e-02	< 3.83e-01	< 3.83e-01	D/A	D/A	D/A	3.83e-01	D/A

Table 1. Data Summary Report.
81-105 (2)

CORE NUMBER: 252
SEGMENT #: 7

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	PPD %	Spk Rec %	Det Limit	Count Err%
S981033004		DSC Exotherm Dry Calculated	Joules/g Dry	98.09	N/A	17.37	15.35	16.36	12.3	N/A	N/A	N/A
S981033004		DSC Exotherm Perkin Elmer	Joules/g	98.09	N/A	8.250	7.290	7.770	12.4	N/A	N/A	N/A
S981033004		Ammonia by ISE Std Additions	ppm	97.74	<0.00e+00	51.50	53.40	52.45	3.62	118.1	1.00e-02	N/A
S981033004		Specific Gravity	g/c.c.	99.14	N/A	1.414	1.432	1.433	2.65	N/A	1.00e-03	N/A
S981033004		A Water by ISE on Perkin Elmer	g/c.c.	98.67	N/A	52.58	52.42	52.50	0.30	N/A	N/A	N/A
S981033004		Leads in Lead on Perkin Elmer	ug/ml	93.89	<9.21e-03	4.78e-02	4.87e-02	4.81e-02	2.49	85.93	1.30e-02	3.47E+01
S981033008		Str-237/240 by TNU-SPEC Resin	ug/ml	101.9	<2.27e-04	1.12e-01	1.15e-01	1.13e-01	0.89	N/A	2.56e-04	2.00E+01
S981033008		Str-237/240 by TNU-SPEC Resin	ug/ml	105.2	<2.99e-05	<3.29e-05	<3.44E-5	<3.44E-5	0.89	N/A	3.29e-05	1.42E+01
S981033008		Uranium ICP-Acid DIL	ug/ml	95.00	<5.00e-02	4.46e+04	4.38e+04	4.42e+04	1.81	51.20	30.10	N/A
S981033008		Boron ICP-Acid DIL	ug/ml	101.2	<5.00e-02	<30.10	<3.01e1	<30.10	N/A	102.0	30.10	N/A
S981033008		Boron ICP-Acid DIL	ug/ml	94.60	<5.00e-02	<30.10	<3.01e1	<30.10	N/A	102.0	30.10	N/A
S981033008		Gallium ICP-Acid DIL	ug/ml	99.60	<1.00e-01	<60.10	<6.01e1	<60.10	N/A	94.60	60.10	N/A
S981033008		Chromium ICP-Acid DIL	ug/ml	103.4	<1.00e-01	<60.10	<6.01e1	<60.10	N/A	94.60	60.10	N/A
S981033008		Lithium ICP-Acid DIL	ug/ml	99.60	<5.00e-02	1.31e+03	1.29e+03	1.30e+03	1.54	96.00	6.010	N/A
S981033008		Potassium ICP-Acid DIL	ug/ml	95.40	<5.00e-02	40.60	39.50	40.05	2.75	93.10	30.10	N/A
S981033008		Manganese ICP-Acid DIL	ug/ml	96.00	<5.00e-02	4.41e+03	4.43e+03	4.42e+03	0.45	101.0	300.0	N/A
S981033008		Sodium ICP-Acid DIL	ug/ml	97.80	<1.00e-01	<6.010	<6.01e0	<6.010	N/A	92.20	6.010	N/A
S981033008		Nickel ICP-Acid DIL	ug/ml	99.40	<2.00e-02	2.03e+05	2.01e+05	2.02e+05	0.99	67.90	60.10	N/A
S981033008		Silicon ICP-Acid DIL	ug/ml	103.2	<5.00e-02	12.00	<1.20e1	12.00	N/A	97.90	12.00	N/A
S981033008		Uranium ICP-Acid DIL	ug/ml	94.60	<1.00e-02	<3.01e2	<3.01e2	<3.01e2	N/A	104.0	30.10	N/A
S981033008		Zinc ICP-Acid DIL	ug/ml	95.60	<1.00e-02	<6.010	<6.01e0	<6.010	N/A	93.00	300.0	N/A
S981033008		Zirconium ICP-Acid DIL	ug/ml	96.20	<1.00e-02	<6.010	<6.01e0	<6.010	N/A	95.40	6.010	N/A
S981033008		Fluoride by Ion Chromatograph	ug/ml	95.39	<2.10e-02	1.02e+03	1.16e+03	1.09e+03	12.8	101.7	319.4	N/A
S981033008		Fluoride-IC-Dionex 4000/4500	ug/ml	95.47	<1.20e-02	<61.81	<6.18e1	<61.81	N/A	83.90	61.81	N/A
S981033008		Chloride-IC-Dionex 4000/4500	ug/ml	100.9	<1.70e-02	5.10e+03	5.44e+03	5.27e+03	6.45	93.88	87.57	N/A
S981033008		Nitrate-IC-Dionex 4000/4500	ug/ml	100.5	<1.28e-01	7.27e+04	7.80e+04	7.54e+04	7.03	97.89	556.3	N/A
S981033008		Bromide by Ion Chromatograph	ug/ml	96.35	<1.00e-01	<6.44e+02	<6.44e2	<6.44e2	N/A	95.56	64.3.9	N/A
S981033008		Nitrate by IC-Dionex 4000/4500	ug/ml	98.71	<1.39e-01	1.88e+05	1.97e+05	1.93e+05	4.68	86.15	716.0	N/A
S981033008		Phosphate-IC-Dionex 4000/4500	ug/ml	99.84	<1.20e-01	8.92e+02	8.73e+02	892.7	2.15	98.17	618.1	N/A
S981033008		Sulfate by IC-Dionex 4000/4500	ug/ml	100.7	<1.38e-01	1.96e+03	2.17e+03	2.07e+03	10.2	99.69	710.8	N/A
S981033008		Oxalate-IC-Dionex 4000/4500	ug/ml	101.1	<1.05e-01	<5.41e+02	<5.41e2	<5.41e2	N/A	101.7	540.9	N/A
S981033008		Cesium-137 by GFA	ug/ml	101.8	9.00e-03	2.15e+02	2.14e+02	2.14e+02	0.47	N/A	N/A	0.0500
S981033012		Strontium (V1) by Spec.	ug/ml	98.22	<2.58e-01	1.09e+03	1.10e+03	1.10e+03	0.91	95.30	1.00e-02	N/A
S981033012		OH- by Pot. Titration	ug/ml	105.0	<1.20e-04	<4.86e0	<4.86e0	<4.86e0	N/A	N/A	4.860	N/A
S981033012		TOC by Persulfate/Coulometry	ug/ml	98.84	<25.00	4.77e+04	5.02e+04	4.90e+04	5.11	N/A	2.50e+03	N/A
S981033012		Uranium-233 by ICP/MS Acid Add	ug/ml	90.67	<5.19e-05	2.55e+03	2.28e+03	2.42e+03	11.2	80.90	40.00	N/A
S981033012		Uranium-234 by ICP/MS Acid Add	ug/ml	N/A	<2.79e-05	<2.101	<2.10e0	<2.10e0	N/A	N/A	2.101	N/A
S981033012		Uranium-235 by ICP/MS Acid Add	ug/ml	102.8	<2.79e-05	<1.128	<1.13e0	<1.13e0	N/A	N/A	1.128	N/A
S981033012		Uranium-236 by ICP/MS Acid Add	ug/ml	N/A	<3.14e-05	<1.271	<1.27e0	<1.27e0	N/A	N/A	1.132	N/A
S981033012		Uranium-238 by ICP/MS Acid Add	ug/ml	102.0	<1.90e-05	<7.68e-1	<7.68e-1	<7.68e-1	N/A	N/A	7.68e-1	N/A
S981033012		Fe 99 by ICP/MS Acid Addition	ug/ml	81.73	<2.00e-06	5.196	5.220	5.203	0.38	82.19	8.10e-02	N/A

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98T003012	D	Sn 126 by ICP/MS Acid Addition	ug/ml										
S98T003012	A	Neptunium-237 by ICP/MS	ug/ml	16.2	<1.50e-04	< 6.075	<6.08e0	n/a	n/a	n/a	6.075	n/a	n/a
S98T003012	A	Plutonium-239 by ICP/MS	ug/ml	14.1	<1.29e-03	< 5.22e-01	<1.06e0	n/a	n/a	129.6	1.058	n/a	n/a
S98T003012	A	Plutonium-240 by ICP/MS	ug/ml	14.1	<1.29e-03	< 5.22e-01	<5.22e-01	n/a	n/a	128.5	5.22e-01	n/a	n/a
S98T003012	A	Plutonium-242 by ICP/MS	ug/ml	158.9	<1.06e-03	< 2.80e-01	<2.80e-01	n/a	n/a	n/a	2.80e-01	n/a	n/a
S98T003012	A	Pu/Am-241 by ICP/MS	ug/ml	158.9	<2.00e-03	< 4.20e-01	<4.20e-01	n/a	n/a	n/a	4.20e-01	n/a	n/a
S98T003012	A	Pu/Am-241 by ICP/MS	ug/ml	158.9	<2.00e-03	< 8.10e-01	<8.10e-01	n/a	n/a	147.7	8.10e-01	n/a	n/a

Table 1. Data Summary Report.
8Y-105 (2)CORE NUMBER: 252
SEGMENT #: 8

SEGMENT PORTION: Drainable Liquid												
Sample#	R #	Analyte	Units	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003005		DSC Exotherm Dry	Joules/g Dry	n/a	n/a	36.63	49.12	42.88	29.1	n/a	n/a	n/a
S98T003005		DSC Exotherm on Perkin Elmer	Joules/g	99.09	n/a	17.54	23.52	20.53	29.1	n/a	n/a	n/a
S98T003005		Ammonia by TSE-Std Additions	Sp.G.	78.98	<10.00	1.19e+02	114.0	116.5	4.29	n/a	10.00	n/a
S98T003005		Specific Gravity	Sp.G.	99.14	n/a	1.454	1.404	1.429	3.50	n/a	1.00e-03	n/a
S98T003005		% Water by TGA on Perkin Elmer	%	98.67	n/a	52.14	52.09	52.12	0.10	n/a	n/a	n/a
S98T003005		Alpha In Liquid Samples	UCI/mL	101.9	<9.21e-03	4.92e-02	6.02e-02	5.47e-02	20.1	n/a	1.30e-02	3.53E+01
S98T003005		Strom-Lum-89/90 High Level	UCI/mL	93.89	<2.27e-04	1.43e-01	1.41e-01	1.42e-01	1.41	n/a	2.53e-04	1.76E+00
S98T003005		Pu-239/240 by TRU-SPEC Resin	UCI/mL	106.2	<2.99e-05	3.48e-05	3.46e-5	n/a	n/a	n/a	3.48e-05	1.00E+02
S98T003009	D	Aluminum-ICP-Acid Dtl.	ug/mL	95.00	<5.00e-02	4.31e+04	4.28e+04	4.30e+04	0.70	n/a	30.10	n/a
S98T003009	D	Boron-ICP-Acid Dtl.	ug/mL	101.2	<5.00e-02	<3.10	<3.01e1	n/a	n/a	n/a	30.10	n/a
S98T003009	D	Barium-ICP-Acid Dtl.	ug/mL	94.60	<1.00e-01	<60.10	<60.1e1	n/a	n/a	n/a	30.10	n/a
S98T003009	D	Bismuth-ICP-Acid Dtl.	ug/mL	99.60	<1.00e-01	<60.10	<60.1e1	n/a	n/a	n/a	60.10	n/a
S98T003009	D	Calcium-ICP-Acid Dtl.	ug/mL	103.4	<1.00e-01	<60.10	<60.1e1	n/a	n/a	n/a	60.10	n/a
S98T003009	D	Chromium-ICP-Acid Dtl.	ug/mL	99.60	<1.00e-02	1.39e+03	1.39e+03	1.39e+03	0.00	n/a	30.10	n/a
S98T003009	D	Iron-ICP-Acid Dtl.	ug/mL	95.40	<5.00e-02	34.00	33.70	33.85	0.89	n/a	300.0	n/a
S98T003009	D	Potassium-ICP-Acid Dtl.	ug/mL	93.20	<5.00e-01	4.38e+03	4.41e+03	4.40e+03	0.68	n/a	6.010	n/a
S98T003009	D	Lithium-ICP-Acid Dtl.	ug/mL	96.00	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003009	D	Manganese-ICP-Acid Dtl.	ug/mL	97.80	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	60.10	n/a
S98T003009	D	Sodium-ICP-Acid Dtl.	ug/mL	95.20	<1.00e-01	2.07e+05	2.04e+05	2.06e+05	1.46	n/a	12.00	n/a
S98T003009	D	Nickel-ICP-Acid Dtl.	ug/mL	92.40	<2.00e-02	<12.00	<1.20e1	n/a	n/a	n/a	30.10	n/a
S98T003009	D	Silicon-ICP-Acid Dtl.	ug/mL	103.2	<5.00e-02	1.71e+02	1.75e+02	1.73e+02	2.31	n/a	30.10	n/a
S98T003009	D	Uranium-ICP-Acid Dtl.	ug/mL	94.60	<5.00e-02	<3.01e+02	<3.01e2	n/a	n/a	n/a	300.0	n/a
S98T003009	D	Zinc-ICP-Acid Dtl.	ug/mL	95.60	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003009	D	Zirconium-ICP-Acid Dtl.	ug/mL	95.20	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T003009		Formate by Ion Chromatograph	ug/mL	95.39	<2.10e-02	1.30e+03	1.36e+03	1.33e+03	4.51	n/a	319.4	n/a
S98T003009		Fluoride-IC-Dionex 4000/4500	ug/mL	95.47	<1.20e-02	<61.81	<6.18e1	n/a	n/a	n/a	61.81	n/a
S98T003009		Chloride-IC-Dionex 4000/4500	ug/mL	100.9	<1.70e-02	5.07e+03	5.14e+03	5.11e+03	1.37	n/a	87.57	n/a
S98T003009		Nitrate by Ion Chromatograph	ug/mL	100.5	<1.08e-01	7.17e+04	7.28e+04	7.22e+04	1.52	n/a	556.3	n/a
S98T003009		Trifluoride by Ion Chromatograph	ug/mL	95.35	<1.25e-01	<6.44e+02	<6.44e2	n/a	n/a	n/a	643.9	n/a
S98T003009		Bromide by Ion Chromatograph	ug/mL	98.71	<1.39e-01	1.94e+05	1.99e+05	1.97e+05	2.54	n/a	716.0	n/a
S98T003009		Nitrate by IC-Dionex 4000/4500	ug/mL	99.84	<1.38e-01	7.28e+02	989.0	858.3	30.4	n/a	618.1	n/a
S98T003009		Phosphate-IC-Dionex 4000/4500	ug/mL	100.7	<1.38e-01	2.13e+03	2.13e+03	2.10e+03	2.38	n/a	710.8	n/a
S98T003009		Sulfate by IC-Dionex 4000/4500	ug/mL	101.1	<1.38e-01	<5.41e+02	<5.41e2	n/a	n/a	n/a	540.9	n/a
S98T003009		Oxalate-IC-Dionex 4000/4500	ug/mL	101.8	9.00e-03	5.23e+02	211.0	212.0	0.94	n/a	n/a	0.0900
S98T003009		Cesium-137 by GEA	ug/mL	98.22	<2.58e-01	1.12e+03	1.13e+03	1.12e+03	0.89	n/a	1.00e-02	n/a
S98T003013		Chromium (VI) by Spec.	ug/mL	105.0	<2.10e-04	<4.860	<4.86e0	n/a	n/a	n/a	4.860	n/a
S98T003013	D	129 by ICP/MS Acid Addition	ug/mL	98.84	<2500.0	1.75e+04	1.75e+04	1.75e+04	2.14	n/a	2.50e+03	n/a
S98T003013		OH- by Pot. Titration	ug/mL	99.67	<5.19e-05	<2.101	<2.10e0	n/a	n/a	n/a	40.00	n/a
S98T003013		TOC by Persulfate/Coulometry	ug/mL	102.8	<2.79e-05	<1.128	<1.13e0	n/a	n/a	n/a	2.101	n/a
S98T003013	D	Uranium-235 by ICP/MS Acid Add	ug/mL	102.8	<2.79e-05	<1.132	<1.13e0	n/a	n/a	n/a	1.128	n/a
S98T003013	D	Uranium-234 by ICP/MS Acid Add	ug/mL	102.8	<2.79e-05	<1.132	<1.13e0	n/a	n/a	n/a	1.132	n/a
S98T003013	D	Uranium-235 by ICP/MS Acid Add	ug/mL	102.0	<1.14e-05	<1.271	<1.27e0	n/a	n/a	n/a	1.271	n/a
S98T003013	D	Uranium-236 by ICP/MS Acid Add	ug/mL	102.0	<1.90e-05	<7.68e-01	<7.68e-01	n/a	n/a	n/a	7.68e-01	n/a
S98T003013	D	Uranium-238 by ICP/MS Acid Add	ug/mL	81.73	<2.00e-06	5.419	5.330	5.374	1.67	n/a	8.10e-02	n/a

Sample#	R /#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003013	D	Sn 126 by ICP/MS	ug/mL	n/a	<1.50e-04	< 6.075	<5.08e0	n/a	n/a	n/a	6.075	n/a
S98T003013	A	Neptunium-237 by ICP/MS	ug/mL	143.3	n/a	< 1.058	<1.06e0	n/a	n/a	n/a	1.058	n/a
S98T003013	A	Plutonium-239 by ICP/MS	ug/mL	141.4	<1.29e-05	<5.22e-01	<5.22e-1	n/a	n/a	n/a	5.22e-01	n/a
S98T003013	A	Plutonium-240 by ICP/MS	ug/mL	n/a	<6.97e-06	<2.80e-01	<2.80e-1	n/a	n/a	n/a	2.80e-01	n/a
S98T003013	A	Plutonium-242 by ICP/MS	ug/mL	n/a	<1.04e-05	<4.19e-01	<4.20e-1	n/a	n/a	n/a	4.20e-01	n/a
S98T003013	A	Pu/Am-241 by ICP/MS	ug/mL	158.9	<2.00e-05	<8.10e-01	<8.10e-1	n/a	n/a	n/a	8.10e-01	n/a

HNF-1663 REV. 0

CHAIN OF CUSTODY FORMS

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 246-01

(2) Supervisor/Sampler M.C. Jones

(3) Tank BY-105

(4) Riser 11B

(5) Cask/Pig Serial No. SV-16-B

(6) Shipment Description:

A. Work Package Number ES-97-00453

B. Cask/Pig Seal Number 3460

C. Date Sample Collected 7-23-98

D. Time Sample Collected 2110

(7) Sampling Data

- Lithium Bromide

Amount N/A

Concentration N/A

- X-Ray

- Peril Sample

- Retrieved Partial Sample Stroke Length 7"

(8) Field Comments:

X-RAY showed 4 1/2" of sample in sampler.

163

(9) Seal Intact Upon Release? ☒ Yes ☐ No

(10) Seal Intact Upon Receipt? ☒ Yes ☐ No

(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6a) ☒ Yes ☐ No

(12) Laboratory Comments:

598T002475

(13) Relinquished By (Sign and PRINT)

Robert J. Paznik

(14) Received By (Sign and PRINT)

Robert J. Paznik

(15) Date/Time 8-7-98

(16) Receiver Comments

(17) Relinquished By (Sign and PRINT)

Robert J. Paznik

(18) Received By (Sign and PRINT)

Robert J. Paznik

(19) Date/Time 8-7-98

(20) Receiver Comments

(21) Relinquished By (Sign and PRINT)

Robert J. Paznik

(22) Received By (Sign and PRINT)

Robert J. Paznik

(23) Date/Time 8-7-98

(24) Receiver Comments

(25) Relinquished By (Sign and PRINT)

Robert J. Paznik

(26) Received By (Sign and PRINT)

Robert J. Paznik

(27) Date/Time

(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>216-02</u>		(2) Supervisor/Sampler <u>M.C. Jones</u>		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank <u>105 BY</u>		(4) Riser <u>113</u>		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Shipment Description: <u>ES-97-00453</u>		(6) Cask/Pig Serial No. <u>C-1055</u>		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6) ☒ Yes ☐ No	
(7) Sampling Date <u>3/6/1</u>		(8) Field Comments: <u>9987002476</u>		(12) Laboratory Comments:	
(13) Relinquished By (Sign and PRINT) <u>Robert J. Spoznik</u>		(14) Received By (Sign and PRINT) <u>Burners</u>		(15) Date/Time <u>10:30 AM 3-7-98</u>	
(16) Relinquished By (Sign and PRINT) <u>Burners</u>		(17) Received By (Sign and PRINT) <u>Shirley King</u>		(18) Date/Time <u>10:55 3-7-98</u>	
(19) Relinquished By (Sign and PRINT) <u>Burners</u>		(20) Received By (Sign and PRINT)		(21) Date/Time	
(22) Relinquished By (Sign and PRINT)		(23) Received By (Sign and PRINT)		(24) Date/Time	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time	
(28) Relinquished By (Sign and PRINT)		(29) Received By (Sign and PRINT)		(30) Date/Time	

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>240-03</u>		(2) Supervisor/Sampler <u>M. C. Jones</u>		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank <u>BX-105</u>		(4) Risk <u>11B</u>		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Shipment Description: <u>ES-97-00163</u>		(6) Cask/Pig Serial No. <u>C-1017</u>		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
(7) Sampling Data		(8) Field Comments:		(12) Laboratory Comments: <u>598T002472</u>	
A. Work Package Number B. Cask/Pig Seal Number C. Date Sample Collected D. Time Sample Collected		Y N ☒ ☐ - Lithium Bromide Amount Concentration - X-Ray - Partial Sample - Retrieved Partial Sample Stroke Length			
(13) Relinquished By (Sign and PRINT) <u>Robert J. Ramirez</u>		(14) Received By (Sign and PRINT) <u>Tom McGinnis</u>		(16) Date/Time <u>8-7-98</u> <u>9:23 AM</u>	
(17) Relinquished By (Sign and PRINT) <u>Robert J. Ramirez</u>		(18) Received By (Sign and PRINT) <u>Tom McGinnis</u>		(20) Date/Time <u>0938</u> <u>8-7-98</u>	
(21) Relinquished By (Sign and PRINT) <u>Robert J. Ramirez</u>		(22) Received By (Sign and PRINT) <u>Tom McGinnis</u>		(24) Date/Time <u>0938</u> <u>8-7-98</u>	
(25) Relinquished By (Sign and PRINT) <u>Robert J. Ramirez</u>		(26) Received By (Sign and PRINT) <u>Tom McGinnis</u>		(28) Date/Time <u>0938</u> <u>8-7-98</u>	

165

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>246-04</u>	(2) Supervisor/Sampler <u>M.C. Jones</u>	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank <u>BY-105</u>	(15) Cask/Pig Serial No. <u>SN-1033</u>	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(6) Shipment Description:		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(7) Sampling Date		(12) Laboratory Comments: <u>S98T002477</u>
A. Work Package Number <u>ES-97-00453</u>	Y ☐ N ☒	
B. Cask/Pig Seal Number <u>3463</u>	Amount <u>N/A</u>	
C. Date Sample Collected <u>7-27-98</u>	Concentration <u>N/A</u>	
D. Time Sample Collected <u>0508</u>	- X-Ray ☐ ☒	
	- Partial Sample ☐ ☒	
	- Retrieved Partial Sample Stroke Length <u>N/A</u>	

(8) Field Comments:

166

(14) Relinquished By (Sign and PRINT) <u>Alvin Roberto S. Parviz</u>	(15) Date/Time <u>10:32 am 8-7-98</u>	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) <u>Alvin Roberto S. Parviz</u>	(18) Date/Time <u>10:55 8-7-98</u>	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) <u>Alvin Roberto S. Parviz</u>	(22) Date/Time <u>10:55 8-7-98</u>	(24) Receiver Comments
(23) Relinquished By (Sign and PRINT) <u>Alvin Roberto S. Parviz</u>	(25) Date/Time <u>10:55 8-7-98</u>	(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 246-00

(2) Surviving Sampler W. Hartley

(3) (a) Riser 11B

(4) Casting Serial No. 1012C

(5) Shipping Description: ES97-00453

(6) Work Package Number 13086

(7) Cask/Pig Seal Number 730-98

(8) Date Sample Collected 1419

(9) Time Sample Collected

(10) Field Comments:

(11) Sampling Data

(12) Lithium Bromide

(13) Amount

(14) Concentration

(15) X-Ray

(16) Partial Sample

(17) Retrieved Partial Sample Stroke Length

(18) Y N

(19) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b)

(20) Laboratory Comments:

(21) Seal Intact Upon Receipt?

(22) Seal Intact Upon Release?

(23) Yes

(24) No

(25) Yes

(26) No

(27) Receiver Comments

(28) Receiver Comments

(29) Receiver Comments

(30) Receiver Comments

(31) Receiver Comments

(32) Receiver Comments

(33) Receiver Comments

(34) Receiver Comments

(35) Receiver Comments

(36) Receiver Comments

(37) Receiver Comments

(38) Receiver Comments

(39) Receiver Comments

(40) Receiver Comments

(41) Receiver Comments

(42) Receiver Comments

2007

CHAIN-OF-CUSTODY RECORD FOR CPO											
(1) Sample Number 270-07	(2) Supervisor/Sampler M. C. Jones	(3) Tank BK-105	(4) Rice 11B	(5) Cask/Pig Serial No. C-2005	(6) Seal Intact Upon Release? ☒ Yes ☐ No	(7) Sampling Data Y ☐ N ☒ Amount N/A Concentration N/A X-Ray ☒ Partial Sample ☒ Retrieved Partial Sample Stroke Length 12" 11" 13"	(8) Field Comments: X-RAY showed 6" of sample in sampler	(9) Seal Intact Upon Receipt? ☒ Yes ☐ No	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6) ☒ Yes ☐ No	(12) Laboratory Comments: 598T002473
(13) Relinquished By (Sign and PRINT) Robert J. Raznik (14) Received By (Sign and PRINT) Tom D. McGinnis						(15) Date/Time 8-2-98	(16) Receiver Comments				
(17) Relinquished By (Sign and PRINT) Tom D. McGinnis (18) Received By (Sign and PRINT) Tom D. McGinnis						(19) Date/Time 09-38	(20) Receiver Comments				
(21) Relinquished By (Sign and PRINT) Tom D. McGinnis (22) Received By (Sign and PRINT) Tom D. McGinnis						(23) Date/Time 8-7-98	(24) Receiver Comments				
(25) Relinquished By (Sign and PRINT) Tom D. McGinnis (26) Received By (Sign and PRINT) Tom D. McGinnis						(27) Date/Time	(28) Receiver Comments				

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number
246-09A

(2) Supervisor/Sampler
WLC Jones

(3) Tank
BY-105

(4) Riser
113

(5) Shipment Description:
ES-97-00453

(6) Cask/Pig Serial No.
G-1009

(7) Sampling Date
7-31-98

(8) Sampling Date
0458

(9) Sampling Date
6"

(10) Sampling Date
6"

(11) Sampling Date
6"

(12) Sampling Date
6"

(13) Sampling Date
6"

(14) Sampling Date
6"

(15) Sampling Date
6"

(16) Sampling Date
6"

(17) Sampling Date
6"

(18) Sampling Date
6"

(19) Sampling Date
6"

(20) Sampling Date
6"

(21) Sampling Date
6"

(22) Sampling Date
6"

(23) Sampling Date
6"

(24) Sampling Date
6"

(25) Sampling Date
6"

(26) Sampling Date
6"

(27) Sampling Date
6"

(28) Sampling Date
6"

(29) Sampling Date
6"

(30) Sampling Date
6"

(31) Sampling Date
6"

(32) Sampling Date
6"

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

Yes ☒ No ☐

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 246A-02R	(2) Supervisor/Sampler M.L. Jones	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank BY-105	(4) Riser 11B	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Shipment Description: ES-97-00453		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(6) Sampling Data		(12) Laboratory Comments:
(7) Sampling Data • Lithium Bromide Amount Concentration • X-Ray • Partial Sample Retrieved Partial Sample Stroke Length		
A. Work Package Number B. Cask/Pig Seal Number C. Date Sample Collected D. Time Sample Collected		
3470 8-3-98 0540		

(8) Field Comments:

170

(13) Relinquished By (Sign and PRINT) Robert J Parvink	(14) Received By (Sign and PRINT) Robert J Parvink	(15) Date/Time 10:23 AM 8-21-98	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) Robert J Parvink	(18) Received By (Sign and PRINT) Robert J Parvink	(19) Date/Time 10:40 8-21-98	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) Robert J Parvink	(22) Received By (Sign and PRINT) Robert J Parvink	(23) Date/Time	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>Q16A-018</u>	(2) Supervisor/Sampler <u>Mike Jones</u>	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank <u>BY-105</u>	(4) Riser <u>11B</u>	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Shipping Description: <u>ES-97-00453</u>		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(6) Work Package Number <u>3471</u>		(12) Laboratory Comments:
(7) Cask/Pig Seal Number <u>8-4-98</u>		
(8) Date Sample Collected <u>0134</u>		
(9) Time Sample Collected		
(10) Field Comments:		

171

(13) Relinquished By (Sign and PRINT) <u>James Seckels</u>	(14) Received By (Sign and PRINT) <u>Mike Jones</u>	(15) Date/Time <u>8-20-98</u>	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) <u>Mike Jones</u>	(18) Received By (Sign and PRINT) <u>Mike Jones</u>	(19) Date/Time <u>8-20-98</u>	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) <u>Mike Jones</u>	(22) Received By (Sign and PRINT) <u>Mike Jones</u>	(23) Date/Time <u>8-20-98</u>	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>2406A-04R</u>		(2) Supervisor/Sampler <u>M.C. JONES</u>		(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank <u>BX-105</u>	(4) Rigar <u>1113</u>	(5) Cask/Pig Serial No. <u>30-G</u>		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(6) Shipment Description:				(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block E & 6a) ☒ Yes
(7) Sampling Date Y N ☒ - Lithium Bromide - Amount - Concentration - X-Ray - Partial Sample - Retrieved Partial Sample Stroke Length				(12) Laboratory Comments:
(8) Field Comments:				

(13) Relinquished By (Sign and PRINT) <u>James Sickels</u>	(14) Received By (Sign and PRINT) <u>BARBARANS</u>	(15) Date/Time <u>1:25 pm</u> <u>8-20-98</u>	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) <u>BARBARANS</u>	(18) Received By (Sign and PRINT) <u>James Sickels</u>	(19) Date/Time <u>14:00</u> <u>8-20-98</u>	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) <u>BARBARANS</u>	(22) Received By (Sign and PRINT) <u>James Sickels</u>	(23) Date/Time	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

172

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 246R-06R		(2) Supervisor/Sampler M.C. Jones		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank 105-BX		(4) Receiver 11B		(10) Seal Intact Upon Receipt? ☐ Yes ☐ No	
(5) Shipment Description: ES-07-00853		(6) Case/Pig Serial No. 881037		(11) Seal Number AND Case/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☐ Yes ☐ No	
(7) Sampling Data		(12) Laboratory Comments:			
A. Work Package Number 3474 B. Case/Pig Seal Number 8-5-98 C. Date Sample Collected 0108 D. Time Sample Collected N/A		(13) Received By (Sign and PRINT) Robert J. Prazich (14) Receiver By (Sign and PRINT) Robert J. Prazich (15) Date/Time 8-21-98 (16) Receiver Comments 1000			
(17) Relinquished By (Sign and PRINT) Robert J. Prazich		(18) Receiver By (Sign and PRINT) Robert J. Prazich			
(19) Relinquished By (Sign and PRINT) Robert J. Prazich		(20) Receiver Comments 1000			
(21) Relinquished By (Sign and PRINT) Robert J. Prazich		(22) Receiver Comments 1000			
(23) Relinquished By (Sign and PRINT) Robert J. Prazich		(24) Receiver Comments 1000			
(25) Relinquished By (Sign and PRINT) Robert J. Prazich		(26) Receiver Comments 1000			
(27) Relinquished By (Sign and PRINT) Robert J. Prazich		(28) Receiver Comments 1000			

CHAIN-OF-CUSTODY RECORD FOR CPO

004

(1) Sample Number

246R-07K

(2) Supervisor/Sampler

M.C. Jones

(3) Tank

BY-105

(4) Risk

11B

(5) Cask/Pig Serial No.

1010-C

(6) Shipment Description:

(7) Sampling Data

Y N

A. Work Package Number

ES-97-00453

B. Cask/Pig Seal Number

3475

C. Data Sample Collected

8-6-98

D. Tens Sample Collected

0334

Lithium Bromide

☒ Y ☐ N

Amount

N/A

Concentration

N/A

X-Ray

☒ Y ☐ N

Partial Sample

☒ Y ☐ N

Retrieved Partial Sample Stroke Length

N/A

(8) Field Comments:

175

(13) Relinquished By (Sign and PRINT)

James Seckels James Seckels

(17) Relinquished By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(12) Relinquished By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(20) Relinquished By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(14) Received By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(18) Received By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(22) Received By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(26) Received By (Sign and PRINT)

D. S. Dunnington D. S. Dunnington

(15) Date/Time

1315

(19) Date/Time

8-20-98

(23) Date/Time

14:00

(27) Date/Time

8-20-98

(16) Receiver Comments

(20) Receiver Comments

(24) Receiver Comments

(28) Receiver Comments

(9) Seal Intact Upon Release?

☒ Yes ☐ No

(10) Seal Intact Upon Receipt?

☒ Yes ☐ No(11) Seal Number AND Cask/Pig SERIAL Number
consistent with this record (Block 6 & 8b)☒ Yes ☐ No

(12) Laboratory Comments:

DISTRIBUTION: Original - Laboratory Copy - Sampling Operations

BC-6001-326 (07/87)

HNF-1663 REV. 0

CHAIN-OF-CUSTODY RECORD FOR CPO

008

(1) Sample Number 240K-08R	(2) Supervisor/Sampler M.C. Jones	(3) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank BY-105	(4) Riser 1103	(9) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Shipment Description: 17-G		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(6) Sampling Data		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 6 & 6b) ☒ Yes ☐ No
(7) Sampling Data		(12) Laboratory Comments:
A. Work Package Number ES-97-00453 B. Cask/Pig Seal Number 3477 C. Date Sample Collected 8-6-98 D. Time Sample Collected 0430		
(8) Field Comments:		
(13) Relinquished By (Sign and PRINT) James Sichel James Sichel	(14) Received By (Sign and PRINT) 2 M.Dunmigh	(15) Date/Time 8-20-98
(17) Relinquished By (Sign and PRINT) 2 M.Dunmigh	(18) Received By (Sign and PRINT) 2 M.Dunmigh	(19) Date/Time 8-20-98
(21) Relinquished By (Sign and PRINT) 2 M.Dunmigh	(22) Received By (Sign and PRINT) 2 M.Dunmigh	(23) Date/Time 8-20-98
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time
(28) Relinquished By (Sign and PRINT)		(29) Received By (Sign and PRINT)

176

11/24/13 01:01 FAX

CHAIN-OF-CUSTODY RECORD FOR CPO

CHAIN-OF-CUSTODY RECORD						☐ Yes ☒ No
1) Sample Number: 240 R-09	(2) Supervisor/Sampler: D. Harris-Hay			(9) Seal Intact Upon Release?		☒ Yes ☐ No
(3) Tank: B405	(4) Riser: 11B	(5) Cask/Pig Serial No.: 1014C	(7) Sampling Data:	(10) Seal Intact Upon Receipt?		☒ Yes ☐ No
(6) Shipment Description: Lithium Bromide				(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b)		☒ Yes ☐ No
A. Work Package Number: ES-99-004530	Amount: N/A			(12) Laboratory Comments:		
B. Cask/Pig Seal Number: 13025	Concentration: N/A					
C. Date Sample Collected: 8-6-98	X-Ray: ☒					
D. Time Sample Collected: 112Z	Partial Sample: ☐					
				- Retrieved Partial Sample Stroke Length:	N/A	

Field Comments:

(13) Relinquished By (Sign and PRINT)	(14) Received By (Sign and PRINT)	(15) Date/Time	(16) Receiver Comments
<i>Edul Robert J. Parvathi</i>	<i>Edul Robert J. Parvathi</i>	10:22 AM 8-21-98	
(17) Relinquished By (Sign and PRINT)	(18) Received By (Sign and PRINT)	(19) Date/Time	(20) Receiver Comments
<i>Edul Robert J. Parvathi</i>	<i>Edul Robert J. Parvathi</i>	10:40 8-21-98	
(21) Relinquished By (Sign and PRINT)	(22) Received By (Sign and PRINT)	(23) Date/Time	(24) Receiver Comments
<i>Edul Robert J. Parvathi</i>	<i>Edul Robert J. Parvathi</i>		
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

BC-6001-326 (07/97)

CHAIN-OF-CUSTODY RECORD FOR CPO

Sample Number <u>250-01</u>	(2) Supervisor Sampler <u>Deborah Delaney</u>	(9) Seal Intact Upon Release? ☒ Yes ☐ No
Tank <u>B4105</u>	(4) Riser <u>7</u>	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
Significant Description: <u>10132</u>		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(7) Sampling Data		(12) Laboratory Comments:
- Lithium Bromide Amount <u>N/A</u>		
Concentration <u>N/A</u>		
- Partial Sample <u>7/12</u>		
- Retrieved Partial Sample Stroke Length		
Field Comments:		
(13) Relinquished By (Sign and PRINT) <u>John Robert J. Parviz</u>		(15) Date/Time <u>8-21-98</u>
(17) Relinquished By (Sign and PRINT) <u>John Robert J. Parviz</u>		(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) <u>John Robert J. Parviz</u>		(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)		(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 150-02		(2) Supervisor/Sampler D. H. H. H.		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Rank Sgt		(4) Rise 7		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Shipments Description ES-98-00452		(6) Cask/Pig Serial No. 10066		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
(7) Sampling Date ES-98-00452		(8) Field Comments: 179		(12) Laboratory Comments:	
A. Work Package Number 13030		Y ☐ N ☒			
B. Cask/Pig Seal Number 8118-98		Amount N/A			
C. Date Sample Collected 1345		Concentration N/A			
D. Time Sample Collected		- X-Ray ☒			
		- Partial Sample ☒			
		- Retrieved Partial Sample Stroke Length 19'			
(13) Relinquished By (Sign and PRINT) James Robert Sparvito		(14) Received By (Sign and PRINT) James Robert Sparvito		(15) Date/Time 1053 8-21-98	
(17) Relinquished By (Sign and PRINT) James Robert Sparvito		(18) Received By (Sign and PRINT) James Robert Sparvito		(19) Date/Time 1046 8-21-98	
(21) Relinquished By (Sign and PRINT) James Robert Sparvito		(22) Received By (Sign and PRINT) James Robert Sparvito		(23) Date/Time	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time	
				(28) Receiver Comments	

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number

250-03

(2) Supervisor Sample

Deborah H. O'Leary

(4) Receipt

BY 105 7

(5) Casting Serial No.

10076

(6) Shipment Description:

ES-97-004520

(7) Sampling Data

Y N

☒ Y

☐ N

Amount

N/A

Concentration

N/A

X-Ray

☒

Partial Sample

☒

Retrieved Partial Sample Stroke Length

19"1

(8) Field Comments:

(9) Seal Intact Upon Release?

☒ Yes

☐ No

(10) Seal Intact Upon Receipt?

☒ Yes

☐ No

(11) Seal Number AND Cast/Pig SERIAL Number consistent with this record? (Block 5 & 6b)

☒ Yes

☐ No

(12) Laboratory Comments:

(13) Relinquished By (Sign and PRINT)

Robert Szank

(14) Received By (Sign and PRINT)

James P. White

(15) Date/Time

1030

(16) Receiver Comments

(17) Relinquished By (Sign and PRINT)

James P. White

(18) Received By (Sign and PRINT)

James P. White

(19) Date/Time

1047

(20) Receiver Comments

(21) Relinquished By (Sign and PRINT)

James P. White

(22) Received By (Sign and PRINT)

James P. White

(23) Date/Time

8-21-98

(24) Receiver Comments

(25) Relinquished By (Sign and PRINT)

(26) Received By (Sign and PRINT)

(27) Date/Time

(28) Receiver Comments

180

HNF-1663 REV. 0

CHAIN-OF-CUSTODY RECORD FOR CPO

11) Sample Number 250-04		12) Supervisor/Sampler MIC. Jones		13) Seal Intact Upon Release? ☒ Yes ☐ No	
13) Tank BSX-105		14) Release 7		15) Seal Intact Upon Receipt? ☒ Yes ☐ No	
16) Shipment Description:		17) Sampling Data		18) Laboratory Comments:	
A. Work Package Number BS-9700452		Y N ☐ Y ☒ N		19) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
B. Cask/Pig Seal Number 3476		Lithium Bromide N/A		20) Laboratory Comments:	
C. Date Sample Collected 8-19-98		Amount N/A			
D. Time Sample Collected 0337		Concentration ☒ X-Ray ☐ Partial Sample			
18) Field Comments:		Retrieved Partial Sample Stroke Length N/A			

113) Relinquished By (Sign and PRINT) James Sickels	114) Received By (Sign and PRINT) Richard Pluchinsky	115) Date/Time 9-8-98	116) Receiver Comments
117) Relinquished By (Sign and PRINT) James Sickels	118) Received By (Sign and PRINT) Richard Pluchinsky	119) Date/Time 9-8-98	120) Receiver Comments
121) Relinquished By (Sign and PRINT)	122) Received By (Sign and PRINT)	123) Date/Time	124) Receiver Comments
125) Relinquished By (Sign and PRINT)	126) Received By (Sign and PRINT)	127) Date/Time	128) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

9-19-98

(1) Sample Number

(2) Super/Isot/Sampler

(3) Tank

(4) Riser

(5) Cask/Pig Serial No.

(6) Shipment Description:

(7) Sampling Date

(8) Work Package Number

(9) Cask/Pig Seal Number

(10) Date Sample Collected

(11) Time Sample Collected

(12) Field Comments:

(13) Relinquished By (Sign and PRINT)

(14) Relinquished By (Sign and PRINT)

(15) Relinquished By (Sign and PRINT)

(16) Relinquished By (Sign and PRINT)

(17) Relinquished By (Sign and PRINT)

(18) Relinquished By (Sign and PRINT)

(19) Relinquished By (Sign and PRINT)

(20) Relinquished By (Sign and PRINT)

(21) Relinquished By (Sign and PRINT)

(22) Relinquished By (Sign and PRINT)

(23) Relinquished By (Sign and PRINT)

(24) Relinquished By (Sign and PRINT)

(25) Relinquished By (Sign and PRINT)

(26) Relinquished By (Sign and PRINT)

(27) Relinquished By (Sign and PRINT)

(28) Relinquished By (Sign and PRINT)

(29) Relinquished By (Sign and PRINT)

(30) Relinquished By (Sign and PRINT)

(31) Relinquished By (Sign and PRINT)

(32) Relinquished By (Sign and PRINT)

(33) Relinquished By (Sign and PRINT)

(34) Relinquished By (Sign and PRINT)

(35) Relinquished By (Sign and PRINT)

(36) Relinquished By (Sign and PRINT)

(37) Relinquished By (Sign and PRINT)

(38) Relinquished By (Sign and PRINT)

(39) Relinquished By (Sign and PRINT)

(40) Relinquished By (Sign and PRINT)

(41) Relinquished By (Sign and PRINT)

(42) Relinquished By (Sign and PRINT)

(43) Relinquished By (Sign and PRINT)

(44) Relinquished By (Sign and PRINT)

(45) Relinquished By (Sign and PRINT)

(46) Relinquished By (Sign and PRINT)

(47) Relinquished By (Sign and PRINT)

Y

N

☐

☒

NA

NA

NA

☐

☒

☐

☒

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

NA

(9) Seal Intact Upon Release?

(10) Seal Intact Upon Receipt?

(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record (Block 9 & 6B)

(12) Laboratory Comments:

(13) Receiver Comments

(14) Receiver Comments

(15) Receiver Comments

(16) Receiver Comments

(17) Receiver Comments

(18) Receiver Comments

(19) Receiver Comments

(20) Receiver Comments

(21) Receiver Comments

(22) Receiver Comments

(23) Receiver Comments

(24) Receiver Comments

(25) Receiver Comments

(26) Receiver Comments

(27) Receiver Comments

(28) Receiver Comments

(29) Receiver Comments

(30) Receiver Comments

(31) Receiver Comments

(32) Receiver Comments

(33) Receiver Comments

(34) Receiver Comments

(35) Receiver Comments

(36) Receiver Comments

(37) Receiver Comments

(38) Receiver Comments

(39) Receiver Comments

(40) Receiver Comments

(41) Receiver Comments

(42) Receiver Comments

(43) Receiver Comments

(44) Receiver Comments

(45) Receiver Comments

(46) Receiver Comments

(47) Receiver Comments

(48) Receiver Comments

(49) Receiver Comments

(50) Receiver Comments

(51) Receiver Comments

(52) Receiver Comments

(53) Receiver Comments

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

No

Yes

OTC Inspection Expires 10/10/98

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 250-06	(2) Supervisor/Sampler <i>Donna D. Davis</i>	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank BY105	(4) Riser 7	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Cask/Pig Serial No. C 1039		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(6) Shipment Description:		(12) Laboratory Comments:
(7) Sampling Data Y N - Lithium Bromide ☐ ☒ <u>N/A</u> Amount <u>N/A</u> Concentration <u>N/A</u> - X-Ray ☐ ☒ - Partial Sample ☐ ☒ - Retrieved Partial Sample Stroke Length <u>N/A</u>		
(8) Field Comments:		
(14) Received By (Sign and PRINT) <i>James D. Kels</i> (15) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i> (16) Received By (Sign and PRINT) <i>James D. Kels</i> (17) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i> (18) Received By (Sign and PRINT) <i>James D. Kels</i> (19) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i> (20) Received By (Sign and PRINT) <i>James D. Kels</i> (21) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i> (22) Received By (Sign and PRINT) <i>James D. Kels</i> (23) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i> (24) Received By (Sign and PRINT) <i>James D. Kels</i> (25) Relinquished By (Sign and PRINT) <i>Donna D. Davis</i>		(15) Date/Time 9-8-98 (16) Receiver Comments (17) Date/Time 9-8-98 (18) Receiver Comments (19) Date/Time 9-8-98 (20) Receiver Comments (21) Date/Time 9-8-98 (22) Receiver Comments (23) Date/Time 9-8-98 (24) Receiver Comments (25) Date/Time 9-8-98 (26) Receiver Comments

183

CHAIN-OF-CUSTODY RECORD FOR CPO

141 Sample Number 250-07		142 Supervisor's Signature <i>Danar Zolman</i>		143 Receiver's Signature <i>Danar Zolman</i>		144 Date 9/8/98	
(3) Tank BX105		(4) Riser 7		(5) Cask/Pig Serial No. C 1053		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(6) Shipment Description:						(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(7) Sampling Data Y N Lithium Bromide ☒ ☐ Amount <u>N/A</u> Concentration <u>N/A</u> X-Ray ☒ ☐ Partial Sample ☒ ☐ Retrieved Partial Sample Stroke Length <u>N/A</u>						(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record (Block 3 & 6) ☒ Yes ☐ No	
(8) Field Comments:						(12) Laboratory Comments:	
120 Relinquished By (Sign and PRINT) <i>Danar Zolman</i> JAMES SICKLES						116 Date/Time 9-8-98	
117 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						118 Date/Time 9-8-98	
121 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						119 Date/Time 9-8-98	
122 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						120 Date/Time 9-8-98	
123 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						121 Date/Time 9-8-98	
124 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						122 Date/Time 9-8-98	
125 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						123 Date/Time 9-8-98	
126 Relinquished By (Sign and PRINT) <i>Danar Zolman</i>						124 Date/Time 9-8-98	

184

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 250-08	(2) Supervisor/Sampler <i>James Sickle</i>	(3) Tank BY 105	(4) Riser 17	(5) Casting Serial No. C 1032	(6) Shipment Description:	(7) Sampling Data - Lithium bromide - Amount - Concentration - X-Ray - Perical Sample - Retrieved Partial Sample Stroke Length	(8) Field Comments: Venting Required As of 9/16/98 Prior To 9-14-98	(9) Seal Intact Upon Release? ☒ Yes ☐ No	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	(11) Seal Number AND Cast/Pig SERIAL Number 598T002850	(12) Laboratory Comments:	(13) Retinquished By (Sign and PRINT) <i>James Sickle</i>	(14) Received By (Sign and PRINT) <i>James Sickle</i>	(15) Date/Time 9-15-98	(16) Receiver Comments
(17) Retinquished By (Sign and PRINT) <i>James Sickle</i>	(18) Received By (Sign and PRINT) <i>James Sickle</i>	(19) Date/Time 9-15-98	(20) Receiver Comments	(21) Retinquished By (Sign and PRINT) <i>James Sickle</i>	(22) Received By (Sign and PRINT) <i>James Sickle</i>	(23) Date/Time 9-15-98	(24) Receiver Comments	(25) Retinquished By (Sign and PRINT) <i>James Sickle</i>	(26) Received By (Sign and PRINT) <i>James Sickle</i>	(27) Date/Time 9-15-98	(28) Receiver Comments				

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 250-09		(2) Supervisor/Sampler James Sackels	
(3) Date 9/10/98	(4) Receiver James Sackels	(5) Cuckling Serial No. 641057	
(6) Shipment Description: C1039			
(7) Sampling Data		Y N	
A. Work Package Number ES-47-00452		☒ ☐	
B. Cuckling Pig Seal Number 13031		N/A	
C. Data Sample Collected 8-21-98		N/A	
D. Time Sample Collected 0915		☒ ☐	
(8) Field Comments:		Retrieved Parcel Sample Stroke Length 134	
(9) Seal Intact Upon Release?		☒ Yes ☐ No	
(10) Seal Intact Upon Receipt?		☒ Yes ☐ No	
(11) User Number AND Cuckling SERIAL Number consistent with this record? (Block 5 & 6)		☒ Yes ☐ No	
(12) Laboratory Comments:			

(13) Relinquished By (Sign and PRINT) James Sackels	(14) Received By (Sign and PRINT) James Sackels	(15) Date/Time 9-8-98	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) James Sackels	(18) Received By (Sign and PRINT) James Sackels	(19) Date/Time 9-8-98	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) James Sackels	(22) Received By (Sign and PRINT) James Sackels	(23) Date/Time 9-8-98	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

186

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 250-09B		(2) Supervisor Sampler D. Gantley		(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank B4185	(4) Reel 7	(5) Casing Serial No. C1034	(6) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(7) Sample Data				(11) Seal Number AND Casing SERIAL Number consistent with this record? (Block 6 & 6b) ☒ Yes ☐ No
A. Work Package Number E5-97-104526				(12) Laboratory Comments:
B. Casing/Fig Seal Number 13032				
C. Date Sample Collected 8-24-98				
D. Time Sample Collected 1100				
(8) Field Comments:				
(13) Relinquished By (Signature and PRINT) James Sickel		(14) Received By (Sign and PRINT) James Sickel		(15) Date/Time 7-8-98
(16) Relinquished By (Sign and PRINT) James Sickel		(17) Received By (Sign and PRINT) James Sickel		(18) Date/Time 7-8-98
(19) Relinquished By (Sign and PRINT) James Sickel		(20) Received By (Sign and PRINT) James Sickel		(21) Date/Time 7-8-98
(22) Relinquished By (Sign and PRINT) James Sickel		(23) Received By (Sign and PRINT) James Sickel		(24) Date/Time 7-8-98
(25) Relinquished By (Sign and PRINT) James Sickel		(26) Received By (Sign and PRINT) James Sickel		(27) Date/Time 7-8-98
(28) Relinquished By (Sign and PRINT) James Sickel		(29) Received By (Sign and PRINT) James Sickel		(30) Date/Time 7-8-98

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <u>251-1</u>		(2) Supervisor/Sampler <u>R. Frank East</u>		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank <u>BV-105</u>		(4) Blaser <u>7</u>		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Ship/Manifest Description: <u>ES-97-00453</u>		(6) Cask/Pig Serial No. <u>28-6</u>		(11) Seal Number AND Cask/Pig Serial Number consistent with this record? (Block 5 & 6a) ☒ Yes ☐ No	
(7) Sampling Data		(12) Laboratory Comments:			
A. Work Package Number <u>ES-97-00453</u>		Y N ☐ Y ☒ N			
B. Cask/Pig Seal Number <u>11376</u>		- Uthium Bromide <u>NA</u>			
C. Date Sample Collected <u>8-24-98</u>		Amount <u>8-24-98</u>			
D. Time Sample Collected <u>1834</u>		Concentration <u>PH</u>			
		- X-Ray ☐ Y ☒ N			
		- Partial Sample ☐ Y ☒ N			
		- Retrieved Partial Sample Stroke Length <u>NA</u>			
(8) Field Comments: push sampler, no x-ray, 800MR dose rate, one drop of liquid on sampler bottom, PH 7 1/2" push of sample 8-24-98					

188

(13) Relinquished By (Sign and PRINT) <u>James Dickel James S. C. 7</u>	(14) Received By (Sign and PRINT) <u>[Signature]</u>	(15) Date/Time <u>9-8-98</u>	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) <u>[Signature]</u>	(18) Received By (Sign and PRINT) <u>[Signature]</u>	(19) Date/Time <u>1495</u>	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) <u>[Signature]</u>	(22) Received By (Sign and PRINT) <u>[Signature]</u>	(23) Date/Time <u>9-8-98</u>	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 251-2		(12) Organization/Computer R Frank East		(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank B4-105	(4) Riser 7	(5) Cask/Pig Serial No. C-2006	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(6) Shipment Description: ES-97-00452				(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No

(7) Sampling Data	Y	N
- Lubrium Bombs	☐	☒
Amount	NA	
Concentration	NA	
- X-ray	☐	☒
- Partial Sample	☐	☒
- Retrieved Partial Sample Stroke Length	NA	

(8) Field Comments:

push sampler, 1R dose rate, no x-ray, sampler has liquid drops on bottom.

189

HNE-1663 REV. 0

(14) Relinquished By (Sign and PRINT) James S. Sackel	(15) Date/Time 9-8-98	(16) Receiver Comments
(17) Relinquished By (Sign and PRINT) James S. Sackel	(18) Date/Time 9-8-98	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) James S. Sackel	(23) Date/Time 9-8-98	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 251-3		(2) Supervisor/Sampler R. Frank Gast		(3) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank BY-105		(4) Riser 7	(5) Cask/Pkg Serial No. 31-G		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(6) Sampling Description UNIDENTIFIED		(7) Sampling Data 31-G		(11) Seal Number AND Cask/Pkg SERIAL Number consistent with this record? (Block E & Gb) ☒ Yes ☐ No	
A. Work Package Number ES-97-00452		V M ☐ ☒		(12) Laboratory Comments	
B. Cask/Pkg Seal Number 11373		- Lithium Bromide NA			
C. Date Sample Collected 8-24-98		Amount NA			
D. Time Sample Collected 2034		Concentration ☒ ☐			
		- X-Ray ☐ ☒			
		- Partial Sample ☐ ☒			
		- Retrieved Partial Sample Stroke Length NA			
(8) Field Comments: Push sampler, 900MK dose rate, no x-ray, sampler has liquid drops on bottom. At 8:20-28					
190					
(13) Relinquished By (Sign and PRINT) James Sickle		(14) Received By (Sign and PRINT) James Sickle		(16) Date/Time 9-8-98	
(17) Relinquished By (Sign and PRINT) James Sickle		(18) Received By (Sign and PRINT) James Sickle		(19) Date/Time 1450	
(21) Relinquished By (Sign and PRINT) James Sickle		(22) Received By (Sign and PRINT) Richard Richards		(23) Date/Time 9-8-98	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time	
				(28) Receiver Comments	

CHAIN-OF-CUSTODY RECORD FOR CPO

11) Barriage Number <u>251-4</u>		14) Supervisor/Sampler <u>R. Frank Frost R. Frank Frost</u>		9) Seal Intact Upon Release? ☒ Yes ☐ No	
13) Tank <u>BY-105</u>		15) Cask/Pig Serial No. <u>C1027</u>		10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
16) Sillment Description <u>7</u>		17) Sampling Date <u>8-24-98</u>		11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (attach 6 & 8b) ☒ Yes ☐ No	
A. Work Package Number <u>ES-97-00451</u>		Y N ☐ Y ☒ N		12) Laboratory Comments:	
B. Cask/Pig Seal Number <u>11374</u>		Lithium Bromide ☐ ☒			
C. Date Sample Collected <u>8-25-98</u>		Amount <u>NA</u>			
D. Time Sample Collected <u>1330</u>		Concentration <u>NA</u>			
		X-Ray ☐ ☒			
		Partial Sample ☐ ☒			
		Retrieved Partial Sample Stock Length <u>NA</u>			
(B) Field Comments: <u>puth samples</u> <u>Dose rate 1 R/hr thru Dir. Spring</u> <u>191</u> <u>OTC-Inspection Expires 10/18/98</u>					
13a) Relinquished By (Sign and PRINT) <u>James J. Kelly</u>		114) Received By (Sign and PRINT) <u>[Signature]</u>		115) Date/Time <u>9-8-98</u>	
13b) Relinquished By (Sign and PRINT) <u>[Signature]</u>		118) Received By (Sign and PRINT) <u>[Signature]</u>		118) Date/Time <u>10-00</u>	
13c) Relinquished By (Sign and PRINT) <u>[Signature]</u>		122) Received By (Sign and PRINT) <u>[Signature]</u>		123) Date/Time <u>9-8-98</u>	
13d) Relinquished By (Sign and PRINT) <u>[Signature]</u>		126) Received By (Sign and PRINT) <u>[Signature]</u>		127) Date/Time <u>[Signature]</u>	
13e) Relinquished By (Sign and PRINT) <u>[Signature]</u>		128) Received By (Sign and PRINT) <u>[Signature]</u>		128) Date/Time <u>[Signature]</u>	

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 251-DYA		(2) Supervisor/Sampler M. G. Jones		(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank BY-105		(4) Riser 7		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Cask/Pig Serial No. SN-5-G		(6) Shipment Description: ES-97-00452		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6a) ☒ Yes ☐ No
(7) Sampling Data		(8) Shipment Description:		(12) Laboratory Comments:
A. Work Package Number 3464 B. Cask/Pig Seal Number 8-22-98 C. Date Sample Collected 1900 D. Time Sample Collected		Y N Lithium Bromide ☒ ☐ Amount N/A Concentration N/A X-Ray ☒ ☐ Partial Sample ☒ ☐ Retrieved Partial Sample Stroke Length N/A		
(13) Field Comments:				
192				
(14) Relinquished By (Sign and PRINT) James Sickle James Sickle	(15) Received By (Sign and PRINT) R. M. Downing	(16) Date/Time 1325-9-8-98	(17) Receiver Comments	
(18) Relinquished By (Sign and PRINT) R. M. Downing	(19) Received By (Sign and PRINT) R. M. Downing	(20) Date/Time 1400-9-8-98	(21) Receiver Comments	
(22) Relinquished By (Sign and PRINT)	(23) Received By (Sign and PRINT)	(24) Date/Time	(25) Receiver Comments	
(26) Relinquished By (Sign and PRINT)	(27) Received By (Sign and PRINT)	(28) Date/Time	(29) Receiver Comments	

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 251-05		(2) Supervisor/Sampler R. FRANK FROST		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank B-105		(4) Blotter 7		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Cask/Pig Serial No. C-2009		(6) Sampling Date ES-97-00452		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record (Block 5 & 6) ☒ Yes ☐ No	
(7) Sampling Date 8-31-98		(8) Field Comments: push samples, 17" sample push, 500 mps down rate, liquid drops on bottom of samples - H		(12) Laboratory Comments:	
A. Work Package Number 11379		Y N ☐ ☒			
B. Cask/Pig Seal Number 8-31-98		Lithium Bromide 8-31-98			
C. Date Sample Collected 1041		Amount NA			
D. Time Sample Collected		Concentration ☒ ☐			
		X-Ray ☒ ☐			
		Partial Sample NA			
		Retrieved Partial Sample Stroke Length NA			
(13) Relinquished By (Sign and PRINT) James S. Sickels		(14) Received By (Sign and PRINT) McDerming		(15) Date/Time 1325 9-8-98	
(16) Relinquished By (Sign and PRINT) McDerming		(17) Received By (Sign and PRINT) McDerming		(18) Date/Time 1500 9-8-98	
(19) Relinquished By (Sign and PRINT) McDerming		(20) Received By (Sign and PRINT) McDerming		(21) Date/Time 9-8-98	
(22) Relinquished By (Sign and PRINT) McDerming		(23) Received By (Sign and PRINT) McDerming		(24) Date/Time 9-8-98	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time	
(28) Relinquished By (Sign and PRINT)		(29) Received By (Sign and PRINT)		(30) Date/Time	

CHAIN-OF-CUSTODY RECORD FOR CPO

CHAIN-OF-CUSTODY RECORD FOR CPO									
(1) Sample Number 252-05	(2) Supervisor/Sampler M.C. Jones	(3) Seal Intact Upon Release? ☒ Yes ☐ No							
(3) Tank BX-105	(4) Riser 7	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No							
(6) Shipman Description: ES-97-00452		(11) Seal Number AND Case/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No							
A. Work Package Number 34581	(7) Sampling Data Y N Lithium Bromide ☒ Amount N/A Concentration N/A X-Ray ☒ Partial Sample ☒ Retrieved Partial Sample Stroke Length N/A	(12) Laboratory Comments:						S98T002731	
B. Case/Pig Seal Number 9-2-98									
C. Date Sample Collected 2030									
D. Time Sample Collected									
(8) Field Comments:									

(14) Relinquished By (Sign and PRINT)	(15) Date/Time	(16) Receiver Comments
James Dickel James Dickel	9-9-98	
(17) Relinquished By (Sign and PRINT)	(18) Date/Time	(20) Receiver Comments
[Signature]	9-9-98	
(21) Relinquished By (Sign and PRINT)	(22) Date/Time	(24) Receiver Comments
[Signature]	9-9-98	
(25) Relinquished By (Sign and PRINT)	(26) Date/Time	(28) Receiver Comments

194

INSTRUMENTATION: Original - Laboratory Copy - Sampling Operations

BC-6001-328 (07/97)

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 852-07		(2) Supervisor/Sampler M.C. Jones		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank BX-105		(4) Rise 7		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Shipment Description: ES-97-00452		(6) Cask/Pig Serial No. 3974 C-2016		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
(7) Sampling Date 9-3-98		(8) Laboratory Comments: 508 T00 2733			
A. Work Package Number 3406		Y N ☐ Y ☒ N			
B. Cask/Pig Seal Number 9-3-98		Lithium Bromide Amount Concentration X-Ray Partial Sample Retrieved Partial Sample Stroke Length N/A			
C. Date Sample Collected 1939					
D. Time Sample Collected					
(13) Relinquished By (Sign and PRINT) James Sicks		(14) Received By (Sign and PRINT) M. Dunnigan		(15) Date/Time 9-9-98	
(17) Relinquished By (Sign and PRINT)		(18) Received By (Sign and PRINT)		(19) Date/Time 9-9-98	
(21) Relinquished By (Sign and PRINT)		(22) Received By (Sign and PRINT)		(23) Date/Time 9-9-98	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Date/Time	

196

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 252-08		(2) Supervisor/Sampler M.C. Jones		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank BY-105		(4) Riser 7		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(5) Shipment Description: ES-97-00452		(6) Cask/Pig Serial No. C-2003		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
(7) Sampling Data		(8) Laboratory Comments: 5987002734		(12) Laboratory Comments:	
A. Work Package Number ES-97-00452 B. Cask/Pig Seal Number 3465 C. Date Sample Collected 9-3-98 D. Time Sample Collected 2022		Y N Lithium Bromide ☐ ☒ Amount N/A Concentration N/A X-Ray ☐ ☒ Partial Sample ☐ ☒ Retrieved Partial Sample Stroke Length N/A			
(8) Field Comments:					
(13) Requisitioned By (Sign and PRINT) James Sickels		(14) Received By (Sign and PRINT) [Signature]		(15) Receiver Comments 9-9-98	
(16) Requisitioned By (Sign and PRINT) [Signature]		(17) Received By (Sign and PRINT) [Signature]		(18) Receiver Comments 9-9-98	
(19) Requisitioned By (Sign and PRINT) [Signature]		(20) Received By (Sign and PRINT) [Signature]		(21) Receiver Comments	
(22) Requisitioned By (Sign and PRINT)		(23) Received By (Sign and PRINT)		(24) Receiver Comments	
(25) Requisitioned By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(27) Receiver Comments	

197

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number 252-9	(12) Supervisor/Sampler R. FARRK Frost	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank By-105	(4) Riser 7	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(6) Siphonant Description: ES-97-00452		(11) Seal Number AND Cast/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(7) Sampling Data		(12) Laboratory Comments:
A. Work Package Number ES-97-00452 B. Cast/Pig Seal Number 11378 C. Date Sample Collected 9-4-98 D. Time Sample Collected 1035 hrs		59ST002735
(17) Sampling Data - Lithium Bromide Y ☐ N ☒ Amount ph 4.48 NA Concentration NA - X-Ray ☐ ☒ - Partial Sample ☐ ☒ - Retrieved Partial Sample Stroke Length NA		
(8) Field Comments: Push sampler, 0.1mg, 25mm dose rate, liquid drops on bottom of sampler. 19th push of sample. No X-ray		
(14) Relinquished By (Sign and PRINT) James Sickle	(14) Received By (Sign and PRINT) M. Dunnington	(16) Date/Time 9-9-98
(17) Relinquished By (Sign and PRINT)	(17) Received By (Sign and PRINT)	(19) Date/Time 9-9-98
(21) Relinquished By (Sign and PRINT)	(21) Received By (Sign and PRINT)	(23) Date/Time 9-9-98
(22) Relinquished By (Sign and PRINT)	(22) Received By (Sign and PRINT)	(27) Date/Time
		(28) Reciever Comments

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number FIELD BLANK	(2) Supervisor/Sampler R. Frank Frost	(9) Seal Intact Upon Release? ☒ Yes ☐ No
(3) Tank BY-105	(4) Risk 11B	(10) Seal Intact Upon Receipt? ☒ Yes ☐ No
(5) Cask/Pig Serial No. sample bottle		(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No
(6) Shipment Description: ES-97-00453		(12) Laboratory Comments: 5987002392
A. Work Package Number 3544	(7) Sampling Data Y N Lithium Bromide ☒ ☐ Amount NA	
B. Cask/Pig Seal Number 7-31-98	Concentration NA	
C. Data Sample Collected 1758	X-Ray ☒ ☐ Partial Sample ☒ ☐	
D. Time Sample Collected	Retrieved Partial Sample Stroke Length NA	

(8) Field Comments:

Field Blank → BY-105, inner #11B
deionized water (Retary) RH 7-31-98

199

2 30 FZ RW

(13) Relinquished By (Sign and PRINT) [Signature]	(14) Received By (Sign and PRINT) [Signature]	(15) Date/Time 0855 8-3-98	(16) Receiver Comments [Signature]
(17) Relinquished By (Sign and PRINT) [Signature]	(18) Received By (Sign and PRINT) [Signature]	(19) Date/Time 11/5 8-3-98	(20) Receiver Comments
(21) Relinquished By (Sign and PRINT) [Signature]	(22) Received By (Sign and PRINT) [Signature]	(23) Date/Time	(24) Receiver Comments
(25) Relinquished By (Sign and PRINT)	(26) Received By (Sign and PRINT)	(27) Date/Time	(28) Receiver Comments

Lithium Bromide sample

CHAIN-OF-CUSTODY RECORD FOR CPO

(1) Sample Number <i>Li Br. sample</i>		(2) Supervisor/Sampler <i>R. Frank Frost</i>		(9) Seal Intact Upon Release? ☒ Yes ☐ No	
(3) Tank <i>BY-105</i>		(4) Right <i>#118</i>		(10) Seal Intact Upon Receipt? ☒ Yes ☐ No	
(6) Shipment Description: <i>sample bath</i>				(11) Seal Number AND Cask/Pig SERIAL Number consistent with this record? (Block 5 & 6b) ☒ Yes ☐ No	
(7) Sampling Data				(12) Laboratory Comments: <i>598002391</i>	
A. Work Package Number <i>ES-97-00453</i>		Y ☒ N ☐			
B. Cask/Pig Seal Number <i>3543</i>		Lithium Bromide <i>100 ml</i>			
C. Date Sample Collected <i>7-31-98</i>		Amount <i>0.3 ml</i>			
D. Time Sample Collected <i>1758 hr</i>		Concentration ☐ X-Ray ☐			
		Partial Sample ☒			
		Retrieved Partial Sample Stroke Length <i>NA</i>			
(8) Field Comments: <i>Lithium Bromide sample - By-105, issue #118</i> <i>(rotary) RH</i> <i>7-31-98</i>					
(13) Relinquished By (Sign and PRINT) <i>John Ebert J. Puznik</i>		(14) Received By (Sign and PRINT) <i>James Knight</i>		(15) Date/Time <i>0855-8-3-98</i>	
(17) Relinquished By (Sign and PRINT) <i>John Ebert J. Puznik</i>		(18) Received By (Sign and PRINT) <i>James Knight</i>		(19) Date/Time <i>1115-8-3-98</i>	
(21) Relinquished By (Sign and PRINT)		(20) Received By (Sign and PRINT)		(23) Date/Time	
(25) Relinquished By (Sign and PRINT)		(26) Received By (Sign and PRINT)		(24) Receiver Comments	
				(28) Receiver Comments	

200

HNF-1663 REV. 0

PHOTOGRAPHS

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

BY 105 Core 246 Seg 2 Riser 11B 8/16/98

BY-105 Core 246 Seg 3 Riser 11B 8/16/98

HNF-1663 REV. 0

BY-105 Core 246 Seg 4 Riser 11B 8/16/98

HNF-1663 REV. 0

BY105 C246 S8 Risey 11B

8/31/88

HNF-1663 REV. 0

BY-105 Core 246

Seg 4

Riser 11B

8/16/98

HNF-1663 REV. 0

BY-105

Core 246

Seg 9A

Riser 11B

8/16/98

BY105 C246R S2R Poser 11B

7/1/98

9/1/98

BY105 C246R 53R Riser 11B

B4105 C246R S4R Riser 11B

7/1/98

8/31/98

B4105 C246R SSR Riser 11B

8/31/68

B4105 C246R 56R Poser 11B

B4105 C446R 57R Fair 11B

9/1/68

B4105 C246F S8K Laser 11B

9/1/98

B4105 C246R S9 Risi 11B

9/1/98

9/2/98

segment 2

core 250

BY-105

7/2/98

B4105 C250 S3 Run 7

7/18/98

B4105 C250 55 Cuiy 7

7/14/98

B4105 C-250 56 Res 7

9/6/68

BY105 C250 S7 Rein 7

9/27/98

R7

Seg. 8

Core 250

BY-105

7/6/68

BY105 C250 S9A Run 7

226

B4105 C251 SI River 7

9/18/98

228

BY-105 Core 251 Segment 2 Pair 7 9/10/98

HNF-1663 REV. 0

9/18/98

R7

Seq. 3

Core 251

BY-105

BY-105 CORE 251 Sep. 4 R7 9/6/68

9/27/98

BY105 C251 549 Run 7

BY105 C251 55 Run 7

9/27/98

BY-105 Core 252 Seg. 5 Riser 7 9/27/98

HNF-1663 REV. 0

BY-105 Core 252 Seg. 6 Riser 7 9/27/98

BY-105 Core 252 Seq. 7 Riser 7 9/27/98

BY-105

Core 252

Sq. 8

Riser 7

9/27/98

BY-105 Core 252 Seq. 9 Riser 7

9/27/98

HNH-1663 REV. 0

SAMPLE HANDLING

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

HNF-1663 REV. 0

CORE COMPOSITE WORKSHEETS

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

Composite Worksheet

Date: 12/16/98
 Tank: BY105

Group # 98000436
 Core 246 Solid Comp (9+9A)

P.C. Stan

Phone # 372-2495

Sample
 ID: 98002397
 Auger/ 209
 Core: 246
 Segment: 9/14

Sample
 ID: 98002588
 Auger/ 209
 Core: 246
 Segment: 9/14

Sample
 ID: 98002588
 Auger/ 209
 Core: 246
 Segment: 9/14

Sample
 ID: 98002588
 Auger/ 209
 Core: 246
 Segment: 9/14

Jar # 15012 (Landing)
 Jar/Vial Size: 60 mL Bin 13
 Initial Weight: 186.09 g
 Final Weight: 63.7 g
 Net Weight: 22.4 g

Jar # 14951 (Landing)
 Jar/Vial Size: 60 mL Bin 13
 Initial Weight: 16.02 g
 Final Weight: 95.4 g
 Net Weight: 20.6 g

Jar # 14951 (Landing)
 Jar/Vial Size: 60 mL
 Initial Weight: g
 Final Weight: g
 Net Weight: g

Jar # 14951 (Landing)
 Jar/Vial Size: 60 mL
 Initial Weight: g
 Final Weight: g
 Net Weight: g

Jar # 14951 (Landing)
 Jar/Vial Size: 60 mL
 Initial Weight: g
 Final Weight: g
 Net Weight: g

Start Time: _____
 End Time: _____

243

Tests: Can 1050 L
1000000
1000000

Jar # 10767
 Jar/Vial Size: 60 mL
 Initial Weight: 128.1 g
 Final Weight: 86.77 g
 Net Weight: 41.1 g

Homogenization
 Time (Min.): _____

Special Instructions:

(after final
 subsample added)

Final Composite Calculation SheetDate: 12-16-98 Technician: YAGGERTotal number of jars used to build the composite: 2Composite Jar Number: 10767 Jar Size: 60

- 1) Sample from segment jar 15012 to Composite jar = 21.6 grams
- 2) Sample from segment jar 14951 to Composite jar = 19.5 grams
- 3) Sample from segment jar _____ to Composite jar = _____ grams
- 4) Sample from segment jar _____ to Composite jar = _____ grams
- 5) Sample from segment jar _____ to Composite jar = _____ grams
- 6) Sample from segment jar _____ to Composite jar = _____ grams
- 7) Sample from segment jar _____ to Composite jar = _____ grams
- 8) Sample from segment jar _____ to Composite jar = _____ grams
- 9) Sample from segment jar _____ to Composite jar = _____ grams
- 10) Sample from segment jar _____ to Composite jar = _____ grams
- 11) Sample from segment jar _____ to Composite jar = _____ grams
- 12) Sample from segment jar _____ to Composite jar = _____ grams
- 13) Sample from segment jar _____ to Composite jar = _____ grams
- 14) Sample from segment jar _____ to Composite jar = _____ grams
- 15) Sample from segment jar _____ to Composite jar = _____ grams
- 16) Sample from segment jar _____ to Composite jar = _____ grams

Total grams collected in Composite Jar # 10767 = 41.1 grams

Composite % Calculation SheetDate: 12-16-98 Technician: YAGER

- 1) Segment Jar #: 15012
21.6 grams from segment jar to composite jar X 100 = 52.6 %
41.1 Total grams in composite jar
- 2) Segment Jar #: 14951
19.5 grams from segment jar to composite jar X 100 = 47.4 %
41.1 Total grams in composite jar
- 3) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 4) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 5) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 6) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 7) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 8) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 9) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar
- 10) Segment Jar #: _____
 _____ grams from segment jar to composite jar X 100 = _____ %
 _____ Total grams in composite jar

Composite Worksheet

Date: 12/16/98
Tank: B4105

Group # 98000436
Core 246 Solid

P.C. Stee
Phone # 372-2495

Sample ID: 98000436
Auger/ Core: 246
Segment: 144

Sample ID: 98000234
Auger/ Core: 246
Segment: 224

Sample ID: 98000256
Auger/ Core: 246
Segment: 444

Sample ID: 98000256
Auger/ Core: 246
Segment: 444

Jar # 14952 (parent)
Jar/Vial Size: 60 mL
Initial Weight: 136.2 g
Final Weight: 111.4 g
Net Weight: 26.2 g

Jar # 14735 (parent)
Jar/Vial Size: 250 mL
Initial Weight: 321.6 g
Final Weight: 295.3 g
Net Weight: 26.3 g

Jar # 15007 (caveat)
Jar/Vial Size: 60 mL
Initial Weight: 168.6 g
Final Weight: 136.2 g
Net Weight: 32.4 g

Jar # 15007 (caveat)
Jar/Vial Size: 60 mL
Initial Weight: 168.6 g
Final Weight: 136.2 g
Net Weight: 32.4 g

246

Start Time: _____
End Time: _____
Tests: Composite
14952
14735
15007

Jar # 14790
Jar/Vial Size: 60 mL
Final Weight: 286.5 g
Initial Weight: 271.4 g
Net Weight: 75.1 g (after final subsample added)

Homogenization Time (Min.): _____

Special Instructions:

Final Composite Calculation SheetDate: 12-16-98 Technician: YAGERTotal number of jars used to build the composite: 3Composite Jar Number: 14790 Jar Size: 250

- 1) Sample from segment jar 14952 to Composite jar = 25.1 grams
- 2) Sample from segment jar 14735 to Composite jar = 25.0 grams
- 3) Sample from segment jar 15007 to Composite jar = 25.0 grams
- 4) Sample from segment jar _____ to Composite jar = _____ grams
- 5) Sample from segment jar _____ to Composite jar = _____ grams
- 6) Sample from segment jar _____ to Composite jar = _____ grams
- 7) Sample from segment jar _____ to Composite jar = _____ grams
- 8) Sample from segment jar _____ to Composite jar = _____ grams
- 9) Sample from segment jar _____ to Composite jar = _____ grams
- 10) Sample from segment jar _____ to Composite jar = _____ grams
- 11) Sample from segment jar _____ to Composite jar = _____ grams
- 12) Sample from segment jar _____ to Composite jar = _____ grams
- 13) Sample from segment jar _____ to Composite jar = _____ grams
- 14) Sample from segment jar _____ to Composite jar = _____ grams
- 15) Sample from segment jar _____ to Composite jar = _____ grams
- 16) Sample from segment jar _____ to Composite jar = _____ grams

Total grams collected in Composite Jar # 14790 = 75.1 grams

Page 1 of 2

Composite % Calculation SheetDate: 12-16-98 Technician: YAGER

- 1) Segment Jar #: 14952
25.1 grams from segment jar to composite jar X 100 = 33.4 %
75.1 Total grams in composite jar
- 2) Segment Jar #: 1735
25.0 grams from segment jar to composite jar X 100 = 33.3 %
75.1 Total grams in composite jar
- 3) Segment Jar #: 15007
25.0 grams from segment jar to composite jar X 100 = 33.3 %
75.1 Total grams in composite jar
- 4) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 5) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 6) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 7) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 8) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 9) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar
- 10) Segment Jar #: _____
_____ grams from segment jar to composite jar X 100 = _____ %
_____ Total grams in composite jar

Composite Worksheet

Date: 12/16/98

Tank: B4105

P.C.: Star

Phone #: 372-2495

109 *48100 2803*

 Samp. ID: 48100 2803
 Aug/Core: 2446R
 Segment: 16 DL

Jar #: 13964 (Archiv)
 Jar Size: 40 mL
 Initial Wgt.: 43.4 g
 Final Wgt.: 32.7 g
 Net Wgt.: 10.7 g

109 *48100 2803*

 Samp. ID: 48100 2803
 Aug/Core: 2446R
 Segment: 7R DL

Jar #: 14760 (parent)
 Jar Size: 250 mL
 Initial Wgt.: 460.3 g
 Final Wgt.: 449.7 g
 Net Wgt.: 10.6 g

109 *48100 2803*

 Samp. ID: 48100 2803
 Aug/Core: 2446R
 Segment: 8R DL

Jar #: 14759 (parent)
 Jar Size: 250 mL
 Initial Wgt.: 462.1 g
 Final Wgt.: 451.9 g
 Net Wgt.: 10.2 g

109 *48100 2803*

 Samp. ID: 48100 2803
 Aug/Core: 2446R
 Segment: 9R DL

Jar #: 13967 (Archiv)
 Jar Size: 40 mL
 Initial Wgt.: 84.8 g
 Final Wgt.: 74.7 g
 Net Wgt.: 10.1 g

249

Samp. ID: 598100 3381 See Parent
 Jar # 13211
 on Page #1

Special Instructions:

Composite Worksheet

Date: 12/16/98

Tank: BY105

Group# 9800437

Core 246LR

P.C.: Star

Phone #: 372-2495

Samp. ID: 48100
Aug/Core: 246LR
Segment: 2R DL

Samp. ID: 48100
Aug/Core: 246LR
Segment: 3R DL

Samp. ID: 48100
Aug/Core: 246LR
Segment: 4R DL

Samp. ID: 48100
Aug/Core: 246LR
Segment: 5R DL

Jar #: 14745 (parent)
Jar Size: 250 mL
Initial Wgt.: 446.7 g
Final Wgt.: 436.1 g
Net Wgt.: 10.6 g

Jar #: 14765 (parent)
Jar Size: 250 mL
Initial Wgt.: 472.0 g
Final Wgt.: 447.5 g
Net Wgt.: 22.5 g

Jar #: 14757 (parent)
Jar Size: 250 mL
Initial Wgt.: 484.0 g
Final Wgt.: 275.9 g
Net Wgt.: 11.1 g

Jar #: 13963 (parent)
Jar Size: 40 mL
Initial Wgt.: 72.9 g
Final Wgt.: 72.7 g
Net Wgt.: 10.2 g

Bin 70

spilled about 10g.

Samp. ID: 48100 3381

Tests: Can't do!

Jar #: 13011
Jar Size: 156+8
Final Wgt.: 6045 mL
Initial Wgt.: 204.1 g

Bin # 136

Initial Wgt.: 837 g / 20.3

Special Instructions: Please make composite using weights listed

Composite % Calculation Sheet

Date: 12/16/48 Technician: R. Hunter

- 1) Segment Jar #: 14745
10.0 grams from segment jar to composite jar X 100 = 12.3 %
81.1 Total grams in composite jar
- 2) Segment Jar #: 14765
10.0 grams from segment jar to composite jar X 100 = 12.3 %
81.1 Total grams in composite jar
- 3) Segment Jar #: 14757
10.0 grams from segment jar to composite jar X 100 = 12.3 %
81.1 Total grams in composite jar
- 4) Segment Jar #: 13963
10.1 grams from segment jar to composite jar X 100 = 12.5 %
81.1 Total grams in composite jar
- 5) Segment Jar #: 13964
10.7 grams from segment jar to composite jar X 100 = 13.2 %
81.1 Total grams in composite jar
- 6) Segment Jar #: 14760
10.2 grams from segment jar to composite jar X 100 = 12.6 %
81.1 Total grams in composite jar
- 7) Segment Jar #: 14759
10.1 grams from segment jar to composite jar X 100 = 12.5 %
81.1 Total grams in composite jar
- 8) Segment Jar #: 13967
10.0 grams from segment jar to composite jar X 100 = 12.3 %
81.1 Total grams in composite jar
- 9) Segment Jar #: N/A
grams from segment jar to composite jar X 100 = N/A %
Total grams in composite jar
- 10) Segment Jar #: N/A
grams from segment jar to composite jar X 100 = N/A %
Total grams in composite jar

Final Composite Calculation Sheet

Date: 12/16/98 Technician: R. HunterTotal number of jars used to build the composite: 8Composite Jar Number: 13011 Jar Size: 125

- 1) Sample from segment jar 14745 to Composite jar = 10.0 grams
- 2) Sample from segment jar 14765 to Composite jar = 10.0 grams
- 3) Sample from segment jar 14757 to Composite jar = 10.0 grams
- 4) Sample from segment jar 13963 to Composite jar = 10.1 grams
- 5) Sample from segment jar 13964 to Composite jar = 10.7 grams
- 6) Sample from segment jar 14760 to Composite jar = 10.2 grams
- 7) Sample from segment jar 14759 to Composite jar = 10.1 grams
- 8) Sample from segment jar 13967 to Composite jar = 10.0 grams
- 9) Sample from segment jar _____ to Composite jar = _____ grams
- 10) Sample from segment jar _____ to Composite jar = _____ grams
- 11) Sample from segment jar _____ to Composite jar = _____ grams
- 12) Sample from segment jar _____ to Composite jar = _____ grams
- 13) Sample from segment jar _____ to Composite jar = _____ grams
- 14) Sample from segment jar _____ to Composite jar = _____ grams
- 15) Sample from segment jar _____ to Composite jar = _____ grams
- 16) Sample from segment jar _____ to Composite jar = _____ grams

Total grams collected in Composite Jar # 13011 = 81.1 grams

Composite Worksheet

Date: 12/16/98

Tank: B4105

Group # 98000490
Core 250 Solid

P.C. Steve

Phone # 372-2495

Jar # 15013 (Bin 131)
Jar/Vial Size: 60 mL
Initial Weight: 158.6 g
Final Weight: 127.5 g
Net Weight: 21.1 g

Jar # 15019 (Bin 131)
Jar/Vial Size: 60 mL
Initial Weight: 118.0 g
Final Weight: 96.8 g
Net Weight: 21.2 g

Jar # 14744 (Quart)
Jar/Vial Size: 250 mL
Initial Weight: 239.1 g
Final Weight: 218.2 g
Net Weight: 20.9 g

Jar # 14738 (Quart)
Jar/Vial Size: 250 mL
Initial Weight: 159.0 g
Final Weight: 136.4 g
Net Weight: 22.6 g

Start Time: _____
End Time: _____

Tests:
Composite
Homogenized
Dry Blend

Jar # 14791
Jar/Vial Size: 250 mL
Final Weight: 249.0 g
Initial Weight: 210.8 g
Net Weight: 38.2 g

Homogenization

Time (Min.): 5 min

Special Instructions:

(after final subsample added)

Composite % Calculation SheetDate: 12/17/98Technician: CM1) Segment Jar #: 15013

$$\frac{20.1 \text{ grams from segment jar to composite jar}}{80.2 \text{ Total grams in composite jar}} \times 100 = \frac{25.1}{} \%$$
2) Segment Jar #: 15019

$$\frac{20.0 \text{ grams from segment jar to composite jar}}{ \text{ Total grams in composite jar}} \times 100 = \frac{24.9}{} \%$$
3) Segment Jar #: 14744

$$\frac{20.2 \text{ grams from segment jar to composite jar}}{ \text{ Total grams in composite jar}} \times 100 = \frac{25.2}{} \%$$
4) Segment Jar #: 15024

$$\frac{19.9 \text{ grams from segment jar to composite jar}}{ \text{ Total grams in composite jar}} \times 100 = \frac{24.8}{} \%$$

5) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

6) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

7) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

8) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

9) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

10) Segment Jar #: _____

$$\frac{\text{grams from segment jar to composite jar}}{\text{Total grams in composite jar}} \times 100 = \%$$

Final Composite Calculation SheetDate: 12/12/98 Technician: CDTotal number of jars used to build the composite: 4Composite Jar Number: 14791 Jar Size: 250

- 1) Sample from segment jar 15013 to Composite jar = 20.1 grams
2) Sample from segment jar 15019 to Composite jar = 20.0 grams
3) Sample from segment jar 14744 to Composite jar = 20.2 grams
4) Sample from segment jar 15024 to Composite jar = 19.9 grams
5) Sample from segment jar _____ to Composite jar = _____ grams
6) Sample from segment jar _____ to Composite jar = _____ grams
7) Sample from segment jar _____ to Composite jar = _____ grams
8) Sample from segment jar _____ to Composite jar = _____ grams
9) Sample from segment jar _____ to Composite jar = _____ grams
10) Sample from segment jar _____ to Composite jar = _____ grams
11) Sample from segment jar _____ to Composite jar = _____ grams
12) Sample from segment jar _____ to Composite jar = _____ grams
13) Sample from segment jar _____ to Composite jar = _____ grams
14) Sample from segment jar _____ to Composite jar = _____ grams
15) Sample from segment jar _____ to Composite jar = _____ grams
16) Sample from segment jar _____ to Composite jar = _____ grams

Total grams collected in Composite Jar # 14791 = 80.2 grams

LABCORE Data Entry Template for Worklist# 25629

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-O

Worklist Comment: BY-105 C246 SEG #1 RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.04</u>	<u>N/A</u>	
98000436	BY-105 (2)	3 SAMPLE	S98T:02475	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000436	BY-105 (2)	4 SAMPLE	S98T:02475	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	5 SAMPLE	S98T:02475	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/ml
98000436	BY-105 (2)	6 SAMPLE	S98T:02475	0	EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	7 SAMPLE	S98T:02475	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	8 SAMPLE	S98T:02475	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>WHC-N-995</u>		
98000436	BY-105 (2)	9 SAMPLE	S98T:02475	0	SLDWT-01	SOLID	<u>N/A</u>	<u>688</u>		g
98000436	BY-105 (2)	10 SAMPLE	S98T:02475	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	11 SAMPLE	S98T:02475	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25629

R.D.H. 8/18/98
Analyst Signature Date

R.D.H. 8/18/98
Analyst Signature Date
8-25-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25630

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-O

Worklist Comment: BY-105 C246 SEG #2 RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	N/A	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.04</u>	N/A	
98000436	BY-105 (2)	3	SAMPLE	S98T:02476	0		DLIQVOL1	SOLID	N/A	<u>0</u>		ml
98000436	BY-105 (2)	4	SAMPLE	S98T:02476	0		DLIQT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	5	SAMPLE	S98T:02476	0		EST.G/ML	SOLID	N/A	<u>0</u>		g/mL
98000436	BY-105 (2)	6	SAMPLE	S98T:02476	0		EXTRUD01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	7	SAMPLE	S98T:02476	0		LLIQT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	8	SAMPLE	S98T:02476	0		NOTEBOOK	SOLID	N/A	<u>WHEN-995</u>		
98000436	BY-105 (2)	9	SAMPLE	S98T:02476	0		SLDWT-01	SOLID	N/A	<u>204.4</u>		g
98000436	BY-105 (2)	10	SAMPLE	S98T:02476	0		APPEAR01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	11	SAMPLE	S98T:02476	0		ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist # 25630

RD/HK 8/18/98
Analyst Signature Date

RD/HK 8/18/98
Analyst Signature Date
8-25-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25631

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-10

Worklist Comment: BY-105 C246 SEG #3 RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.04</u>	N/A	
98000436	BY-105 (2)	3 SAMPLE	S98T:02472	0	DLIQVOL1	SOLID	N/A	<u>0</u>		ml
98000436	BY-105 (2)	4 SAMPLE	S98T:02472	0	DLIQWT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	5 SAMPLE	S98T:02472	0	EST.G/ML	SOLID	N/A	<u>0</u>		g/ml
98000436	BY-105 (2)	6 SAMPLE	S98T:02472	0	EXTRUD01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	7 SAMPLE	S98T:02472	0	LLIQWT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	8 SAMPLE	S98T:02472	0	NOTEBOOK	SOLID	N/A	<u>WAC-12995</u>		
98000436	BY-105 (2)	9 SAMPLE	S98T:02472	0	SLDWT-01	SOLID	N/A	<u>14.6</u>		g
98000436	BY-105 (2)	10 SAMPLE	S98T:02472	0	APPEAR01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	11 SAMPLE	S98T:02472	0	ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist # 25631

R.D. Hale 8/18/98
Analyst Signature Date

R.D. Hale 8/18/98
Analyst Signature Date
ANG C2 8-25-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25632

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246 SEG #4 RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMP. E#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>50000</u>	<u>500.04</u>	<u>N/A</u>	
98000436	BY-105 (2)	3	SAMPLE	S98T102477	0		DL1QVOL1	SOLID	<u>N/A</u>	<u>0</u>		mL
98000436	BY-105 (2)	4	SAMPLE	S98T102477	0		DL1QWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	5	SAMPLE	S98T102477	0		EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/mL
98000436	BY-105 (2)	6	SAMPLE	S98T102477	0		EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	7	SAMPLE	S98T102477	0		LL1QWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	8	SAMPLE	S98T102477	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>WHEN 995</u>		
98000436	BY-105 (2)	9	SAMPLE	S98T102477	0		SLDWT-01	SOLID	<u>N/A</u>	<u>104.1</u>		g
98000436	BY-105 (2)	10	SAMPLE	S98T102477	0		APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	11	SAMPLE	S98T102477	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 25632

RD HLB 8/18/98
Analyst Signature Date

RD HLB 8/18/98
Analyst Signature Date
AN 8 D. W. 8.25.98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25633

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246 SEG #9 RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>50.00</u>	<u>50.04</u>	<u>N/A</u>	
98000436	BY-105 (2)	3 SAMPLE	S98T012473	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000436	BY-105 (2)	4 SAMPLE	S98T012473	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	5 SAMPLE	S98T012473	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/mL
98000436	BY-105 (2)	6 SAMPLE	S98T012473	0	EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	7 SAMPLE	S98T012473	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000436	BY-105 (2)	8 SAMPLE	S98T012473	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>WJH-U-95</u>		
98000436	BY-105 (2)	9 SAMPLE	S98T012473	0	SLDWT-01	SOLID	<u>N/A</u>	<u>139.7</u>		g
98000436	BY-105 (2)	10 SAMPLE	S98T012473	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000436	BY-105 (2)	11 SAMPLE	S98T012473	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25633

RJH 8/18/98
Analyst Signature Date

RJH 8/18/98
Analyst Signature Date
8-25-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25634

Analyst: _____ Instrument: BA000 _____ Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246 SEG #9A RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE:#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>2000</u>	N/A	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.04</u>	<u>500.04</u>	N/A	
98000436	BY-105 (2)	3	SAMPLE	S9BTC02474	0		DLIQVOL1	SOLID	N/A	<u>0</u>		ml
98000436	BY-105 (2)	4	SAMPLE	S9BTC02474	0		DLIQWT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	5	SAMPLE	S9BTC02474	0		EST.G/ML	SOLID	N/A	<u>0</u>		g/ml
98000436	BY-105 (2)	6	SAMPLE	S9BTC02474	0		EXTRUD01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	7	SAMPLE	S9BTC02474	0		LLIQWT01	SOLID	N/A	<u>0</u>		g
98000436	BY-105 (2)	8	SAMPLE	S9BTC02474	0		NOTEBOOK	SOLID	N/A	<u>6.11C-1-995</u>		
98000436	BY-105 (2)	9	SAMPLE	S9BTC02474	0		SLDWT-01	SOLID	N/A	<u>42.3</u>		g
98000436	BY-105 (2)	10	SAMPLE	S9BTC02474	0		APPEAR01	SOLID	N/A	<u>COMPLETE</u>		
98000436	BY-105 (2)	11	SAMPLE	S9BTC02474	0		ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist # 25634

RDH/b 8/18/98
Analyst Signature Date

Co Hale 8/18/98
Analyst Signature Date

And Du 8.25.98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25798

Analyst: SOH drg ^{27 9-3-98} Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246 SEG.#8 RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000436	BY-105 (2)	3 SAMPLE	S98T001625	0	DL1QVOL1	SOLID	<u>N/A</u>	<u>Ø</u>		ml
98000436	BY-105 (2)	4 SAMPLE	S98T001625	0	DL1QWT01	SOLID	<u>N/A</u>	<u>Ø</u>		g
98000436	BY-105 (2)	5 SAMPLE	S98T001625	0	EST.G/ML	SOLID	<u>N/A</u>	<u>Ø</u>		g/ml
98000436	BY-105 (2)	6 SAMPLE	S98T001625	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000436	BY-105 (2)	7 SAMPLE	S98T001625	0	LL1QWT01	SOLID	<u>N/A</u>	<u>Ø</u>		g
98000436	BY-105 (2)	8 SAMPLE	S98T001625	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>11-995</u>		
98000436	BY-105 (2)	9 SAMPLE	S98T001625	0	SLDWT-01	SOLID	<u>N/A</u>	<u>42.6</u>		g
98000436	BY-105 (2)	10 SAMPLE	S98T001625	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000436	BY-105 (2)	11 SAMPLE	S98T001625	0	ORGVOL01	SOLID	<u>N/A</u>	<u>Ø</u>		ml

Final page for worklist # 25798

SOH drg ⁹⁻³⁻⁹⁸
Analyst Signature Date

SOH drg ⁹⁻³⁻⁹⁸
Analyst Signature Date
ASD ⁹⁻⁴⁻⁹⁸

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

08/24/98 10:14

LABCORE Data Entry Template for Worklist# 25803

Analyst: W. J. R. 9-3-98 Instrument: BA000 Book # N/AMethod: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246F SEG.#2R RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2	INSTCHK02		EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3	SAMPLE	S98T102626	0	DLTQVOL1	SOLID	<u>N/A</u>	<u>250</u>	ml
98000437	BY-105 (2)	4	SAMPLE	S98T102626	0	DLTQWT01	SOLID	<u>N/A</u>	<u>356.6</u>	g
								<u>1.43</u>		
98000437	BY-105 (2)	5	SAMPLE	S98T102626	0	EST.G/ML	SOLID	<u>N/A</u>	<u>1.43</u>	g/mL
98000437	BY-105 (2)	6	SAMPLE	S98T102626	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>	
98000437	BY-105 (2)	7	SAMPLE	S98T102626	0	LLTQWT01	SOLID	<u>N/A</u>	<u>0</u>	g
98000437	BY-105 (2)	8	SAMPLE	S98T102626	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N995</u>	
98000437	BY-105 (2)	9	SAMPLE	S98T102626	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>	g
98000437	BY-105 (2)	10	SAMPLE	S98T102626	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
98000437	BY-105 (2)	11	SAMPLE	S98T102626	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>	ml

Final page for worklist # 25803

Analyst Signature W. J. R. 9-3-98 DateAnalyst Signature W. J. R. 9-3-98 DateAnalyst Signature W. J. R. 9-4-98 Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25805

Analyst: By 9-3-98
For J. J. Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY--105 C24CR SEG.#3R RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMP. #	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3 SAMPLE	S98TJ02627	0	DLTQV01	SOLID	<u>N/A</u>	<u>250</u>		ml
98000437	BY-105 (2)	4 SAMPLE	S98TJ02627	0	DLTQW01	SOLID	<u>N/A</u>	<u>363.4</u>		g
98000437	BY-105 (2)	5 SAMPLE	S98TJ02627	0	EST, G/ML	SOLID	<u>N/A</u>	<u>645</u>		g/ml
98000437	BY-105 (2)	6 SAMPLE	S98TJ02627	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	7 SAMPLE	S98TJ02627	0	LLTQW01	SOLID	<u>N/A</u>	<u>Ø</u>		g
98000437	BY-105 (2)	8 SAMPLE	S98TJ02627	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000437	BY-105 (2)	9 SAMPLE	S98TJ02627	0	SLDWT-01	SOLID	<u>N/A</u>	<u>Ø</u>		g
98000437	BY-105 (2)	10 SAMPLE	S98TJ02627	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	11 SAMPLE	S98TJ02627	0	ORGVOL01	SOLID	<u>N/A</u>	<u>Ø</u>		ml

Final page for worklist # 25805

D. J. J. 9-3-98
Analyst Signature Date

D. J. J. 9-3-98
Analyst Signature Date
And J. J. 9-4-98

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 25807

Analyst: DR Jha 9-3-98 Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246R SEG.#4R RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>22.50</u>	<u>20.50</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3	SAMPLE	S981002628	0		DLIQVOL1	SOLID	<u>N/A</u>	<u>250</u>		ml
98000437	BY-105 (2)	4	SAMPLE	S981002628	0		DLIQWT01	SOLID	<u>N/A</u>	<u>334.2</u>		g
98000437	BY-105 (2)	5	SAMPLE	S981002628	0		EST.G/ML	SOLID	<u>N/A</u>	<u>1.34</u>		g/ml
98000437	BY-105 (2)	6	SAMPLE	S981002628	0		EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	7	SAMPLE	S981002628	0		LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	8	SAMPLE	S981002628	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000437	BY-105 (2)	9	SAMPLE	S981002628	0		SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	10	SAMPLE	S981002628	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	11	SAMPLE	S981002628	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25807

DR Jha 9-3-98
Analyst Signature Date

DR Jha 9-3-98
Analyst Signature Date
ASD 9-4-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25808

Analyst: 19204 204 Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C24CR SEG.#5R RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMP.E#	R A	TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3 SAMPLE	S981102629	0	DL1QVOL1	SOLID	<u>N/A</u>	<u>65</u>		ml
98000437	BY-105 (2)	4 SAMPLE	S981102629	0	DL1QWT01	SOLID	<u>N/A</u>	<u>114.8</u>		g
98000437	BY-105 (2)	5 SAMPLE	S981102629	0	EST.G/ML	SOLID	<u>N/A</u>	<u>627</u>		g/ml
98000437	BY-105 (2)	6 SAMPLE	S981102629	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	7 SAMPLE	S981102629	0	LL1QWT01	SOLID	<u>N/A</u>	<u>< 5 did not</u>	<u>Rebin</u>	
98000437	BY-105 (2)	8 SAMPLE	S981102629	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-985</u>		
98000437	BY-105 (2)	9 SAMPLE	S981102629	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	10 SAMPLE	S981102629	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	11 SAMPLE	S981102629	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25808

David R. Juh 9-3-98
Analyst Signature Date

David R. Juh 9-3-98
Analyst Signature Date

Ant
9-4-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25809

Analyst: DP 9-3-98
Rott Dr. Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-D

Worklist Comment: BY-105 C246R SEG.#6R RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20-00</u>	<u>20.750</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3	SAMPLE	S98T102630	0		DLIQVOL1	SOLID	<u>N/A</u>	<u>300</u>		ml
98000437	BY-105 (2)	4	SAMPLE	S98T102630	0		DLIQWT01	SOLID	<u>N/A</u>	<u>374.2</u>		g
98000437	BY-105 (2)	5	SAMPLE	S98T102630	0		EST.G/ML	SOLID	<u>N/A</u>	<u>1.25</u>		g/ml
98000437	BY-105 (2)	6	SAMPLE	S98T102630	0		EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	7	SAMPLE	S98T102630	0		LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	8	SAMPLE	S98T102630	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000437	BY-105 (2)	9	SAMPLE	S98T102630	0		SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	10	SAMPLE	S98T102630	0		APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	11	SAMPLE	S98T102630	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25809

DP Rott 9-3-98
Analyst Signature Date

DP Rott 9-3-98
Analyst Signature Date
AD 9-7-98

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 25811

Analyst: 7-3-98 Drj Instrument: BA000 Book # n/a

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246R SEG.#7R RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3 SAMPLE	S98'002631	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>2.50</u>		ml
98000437	BY-105 (2)	4 SAMPLE	S98'002631	0	DLIQWT01	SOLID	<u>N/A</u>	<u>354.2</u>		g
98000437	BY-105 (2)	5 SAMPLE	S98'002631	0	EST.G/ML	SOLID	<u>N/A</u>	<u>1.42</u>		g/ml
98000437	BY-105 (2)	6 SAMPLE	S98'002631	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	7 SAMPLE	S98'002631	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	8 SAMPLE	S98'002631	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>n-995</u>		
98000437	BY-105 (2)	9 SAMPLE	S98'002631	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	10 SAMPLE	S98'002631	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	11 SAMPLE	S98'002631	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25811

Drj Analyst Signature Date 9-3-98

Drj Analyst Signature Date 9-3-98
Aug 21 9-4-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25814

Analyst: Dr. J. R. Smith Instrument: BA000 Book # 1/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C246F: SEG.#8R RISER #11B EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3	SAMPLE	S98T102632	0		DLTQVOL1	SOLID	<u>N/A</u>	<u>250</u>		ml
98000437	BY-105 (2)	4	SAMPLE	S98T102632	0		DLTQWT01	SOLID	<u>N/A</u>	<u>348.5</u>		g
98000437	BY-105 (2)	5	SAMPLE	S98T102632	0		EST.G/ML	SOLID	<u>N/A</u>	<u>1.39</u>		g/ml
98000437	BY-105 (2)	6	SAMPLE	S98T102632	0		EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	7	SAMPLE	S98T102632	0		LLTQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	8	SAMPLE	S98T102632	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000437	BY-105 (2)	9	SAMPLE	S98T102632	0		SLDWT-01	SOLID	<u>N/A</u>	<u>92.1</u>		g
98000437	BY-105 (2)	10	SAMPLE	S98T102632	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000437	BY-105 (2)	11	SAMPLE	S98T102632	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25814

Dr. J. R. Smith 9-3-98
Analyst Signature Date

Dr. J. R. Smith 9-3-98
Analyst Signature Date

Dr. J. R. Smith 9-4-98

Data Entry Comments:

worklistrpt Version 2.1 05/15/95
08/24/98 13:53

Page: 1

LABCORE Data Entry Template for Worklist# 25815

Analyst: 293-398
LOT drj Instrument: BA000 Book # 11/A

Method: LO-160-103 Rev/Mod D-5

Worklist Comment: BY-105 C246R SEG.#9 RISER #11B EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000437	BY-105 (2)	3 SAMPLE	S98'002633	0	DLIGHT01	SOLID	<u>N/A</u>	<u>105</u>		ml
98000437	BY-105 (2)	4 SAMPLE	S98'002633	0	DLIGHT01	SOLID	<u>N/A</u>	<u>151.2</u>		g
98000437	BY-105 (2)	5 SAMPLE	S98'002633	0	EST.G/ML	SOLID	<u>N/A</u>	<u>1.44</u>		g/ml
98000437	BY-105 (2)	6 SAMPLE	S98'002633	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	7 SAMPLE	S98'002633	0	LLIGHT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000437	BY-105 (2)	8 SAMPLE	S98'002633	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>1-995</u>		
98000437	BY-105 (2)	9 SAMPLE	S98'002633	0	SLDWT-01	SOLID	<u>N/A</u>	<u>275.0</u>		g
98000437	BY-105 (2)	10 SAMPLE	S98'002633	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000437	BY-105 (2)	11 SAMPLE	S98'002633	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25815

293-398
Analyst Signature Date

293-398
Analyst Signature Date

294-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 25830

Analyst: W. R. H. Jr. Instrument: BA000 Book # 11/A

Method: LO-160-103 Rev/Mod P-D

Worklist Comment: BY-105 C250 SEG.#1 RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>50.00</u>	<u>50.0</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T102634	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000490	BY-105 (2)	4 SAMPLE	S98T102634	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T102634	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>	<u>9-3-98</u>	ml
98000490	BY-105 (2)	6 SAMPLE	S98T102634	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T102634	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	8 SAMPLE	S98T102634	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>11-995</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T102634	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T102634	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T102634	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25830

W. R. H. Jr. 9-3-98
Analyst Signature Date

W. R. H. Jr. 9-3-98
Analyst Signature Date
9-4-98

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 25831

Analyst: drj _____ Instrument: BA000 _____ Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C250 SEG.#2 RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMP.E#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.00</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T102635	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000490	BY-105 (2)	4 SAMPLE	S98T102635	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T102635	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/ml
98000490	BY-105 (2)	6 SAMPLE	S98T102635	0	EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T102635	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	8 SAMPLE	S98T102635	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T102635	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T102635	0	APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T102635	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 25831

D. R. Jones 9-4-98
Analyst Signature Date

D. R. Jones 9-4-98
Analyst Signature Date
Ant Du 9.4.98

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 25832

Analyst: Dr. J. R. Ahl Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-O

Worklist Comment: BY-105 C250 BEG.#3 RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01		EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02		EXTRUD01	SOLID	<u>500.00</u>	<u>500.0</u>	<u>N/A</u>	
98000490	BY-105 (2)	3	SAMPLE	S98T:02636	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>	mL
98000490	BY-105 (2)	4	SAMPLE	S98T:02636	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>	g
98000490	BY-105 (2)	5	SAMPLE	S98T:02636	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>	g/mL
98000490	BY-105 (2)	6	SAMPLE	S98T:02636	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>	
98000490	BY-105 (2)	7	SAMPLE	S98T:02636	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>	g
98000490	BY-105 (2)	8	SAMPLE	S98T:02636	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>	
98000490	BY-105 (2)	9	SAMPLE	S98T:02636	0	SLDWT-01	SOLID	<u>N/A</u>	<u>92.9</u>	g
98000490	BY-105 (2)	10	SAMPLE	S98T:02636	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>	
98000490	BY-105 (2)	11	SAMPLE	S98T:02636	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>	mL

Final page for worklist # 25832

Dr. J. R. Ahl 9-3-98
Analyst Signature Date

Dr. J. R. Ahl 9-3-98
Analyst Signature Date

Dr. J. R. Ahl 9-4-98

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 26100

Analyst: ABC Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 Core #50 Seg #4 Riser #7 - Extrusion

GROUP	PROJECT	S TYPE	SAMPLE #	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.00</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T102725	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		mL
98000490	BY-105 (2)	4 SAMPLE	S98T102725	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T102725	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/mL
98000490	BY-105 (2)	6 SAMPLE	S98T102725	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T102725	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	8 SAMPLE	S98T102725	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-995</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T102725	0	SLDWT-01	SOLID	<u>N/A</u>	<u>58.7</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T102725	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T102725	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 26100

ABC Complete 9/14/98
Analyst Signature Date

Alan Blomberg 9/14/98
Analyst Signature Date

ABC 9.17.98

Data Entry Comments:
Fixed 9-17-98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number.
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26102

Analyst: DRJ Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 core 250 SEG.#5 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE #	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>2000</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T102726	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>2</u>		ml
98000490	BY-105 (2)	4 SAMPLE	S98T102726	0	DLIQWT01	SOLID	<u>N/A</u>	<u>2</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T102726	0	EST.G/ML	SOLID	<u>N/A</u>	<u>2</u>		g/ml
98000490	BY-105 (2)	6 SAMPLE	S98T102726	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T102726	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	8 SAMPLE	S98T102726	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>N-99.5</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T102726	0	SLDWT-01	SOLID	<u>N/A</u>	<u>122.6</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T102726	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T102726	0	ORGVOL01	SOLID	<u>N/A</u>	<u>2</u>		ml

Final page for worklist # 26102

Alan Blanchard 9/17/98
Analyst Signature Date

DRJ 9-16-98
Analyst Signature Date
AE 9-17-98

Data Entry Comments:

Fixed 9-17-98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26116

Analyst: ABC Instrument: BA000 Book # 2114

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C250 REG.#6 RISER #7 EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	20.00	20.00	N/A	
		2	INSTCHK02				EXTRUD01	SOLID	500.00	500.10	N/A	
98000490	BY-105 (2)	3	SAMPLE	S98TC02727	0		DLIQVOL1	SOLID	N/A	2		ml
98000490	BY-105 (2)	4	SAMPLE	S98TC02727	0		DLIQWT01	SOLID	N/A	2		g
98000490	BY-105 (2)	5	SAMPLE	S98TC02727	0		EST.G/ML	SOLID	N/A	2		g/ml
98000490	BY-105 (2)	6	SAMPLE	S98TC02727	0		EXTRUD01	SOLID	N/A	Complete		
98000490	BY-105 (2)	7	SAMPLE	S98TC02727	0		LLIQWT01	SOLID	N/A	2		g
98000490	BY-105 (2)	8	SAMPLE	S98TC02727	0		NOTEBOOK	SOLID	N/A	1-995		
98000490	BY-105 (2)	9	SAMPLE	S98TC02727	0		SLDWT-01	SOLID	N/A	157.2		g
98000490	BY-105 (2)	10	SAMPLE	S98TC02727	0		APPEAR01	SOLID	N/A	Complete		
98000490	BY-105 (2)	11	SAMPLE	S98TC02727	0		ORGVOL01	SOLID	N/A	2		ml

Final page for worklist # 26116

Alan Campbell 9/14/98
Analyst Signature Date

Alan Campbell 9/14/98
Analyst Signature Date
ABC 9/17/98

Data Entry Comments:

Fixed 9-17-98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26117

Analyst: abc Instrument: BA000 Book # NAMethod: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C250 SEG.#7 RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>2000</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T002728	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000490	BY-105 (2)	4 SAMPLE	S98T002728	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T002728	0	EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/ml
98000490	BY-105 (2)	6 SAMPLE	S98T002728	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T002728	0	LLIQWT01	SOLID	<u>N/A</u>	<u>KS</u>	<u>did not</u>	<u>RETAIN</u>
98000490	BY-105 (2)	8 SAMPLE	S98T002728	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>na</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T002728	0	SLDWT-01	SOLID	<u>N/A</u>	<u>60</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T002728	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T002728	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 26117

Analyst Signature Alan B Campbell Date 9/18/98Analyst Signature R. Huh Date 9/23/98Approved R. Schwach 9/23/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

277

LABCORE Data Entry Template for Worklist# 26118

Analyst: R. Hale Instrument: BA000 Book # NAMethod: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C250 SEG. #9A RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACT. AL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>50.00</u>	<u>50.10</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T002729	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000490	BY-105 (2)	4 SAMPLE	S98T002729	0	DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T002729	0	EST. G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/mL
98000490	BY-105 (2)	6 SAMPLE	S98T002729	0	EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T002729	0	LLIQWT01	SOLID	<u>N/A</u>	<u>25</u>	<u>DID NOT RETAIN</u>	g
98000490	BY-105 (2)	8 SAMPLE	S98T002729	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>NA</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T002729	0	SLDWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T002729	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T002729	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 26118

Analyst Signature R. Hale Date 9/22/98Analyst Signature R. Hale Date 9/22/98Approved R. Hale 9/25/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

278

LABCORE Data Entry Template for Worklist# 26119

Analyst: Rost rdh Instrument: BA000 Book # NA

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C250 SEG. #9B RISER #7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>2.100</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>5.000</u>	<u>500.0</u>	<u>N/A</u>	
98000490	BY-105 (2)	3 SAMPLE	S98T002730	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>160</u>		mL
98000490	BY-105 (2)	4 SAMPLE	S98T002730	0	DLIQWT01	SOLID	<u>N/A</u>	<u>210.2</u>		g
98000490	BY-105 (2)	5 SAMPLE	S98T002730	0	EST.G/ML	SOLID	<u>N/A</u>	<u>1.31</u>		g/mL
98000490	BY-105 (2)	6 SAMPLE	S98T002730	0	EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000490	BY-105 (2)	7 SAMPLE	S98T002730	0	LLIQWT01	SOLID	<u>N/A</u>	<u><5</u>	<u>210 not 150</u>	<u>g</u>
98000490	BY-105 (2)	8 SAMPLE	S98T002730	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>na</u>		
98000490	BY-105 (2)	9 SAMPLE	S98T002730	0	SLDWT-01	SOLID	<u>N/A</u>	<u>40.1</u>		g
98000490	BY-105 (2)	10 SAMPLE	S98T002730	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000490	BY-105 (2)	11 SAMPLE	S98T002730	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 26119

R. H. H. 9/22/98
Analyst Signature Date

R. H. H. 9/22/98
Analyst Signature Date

Approved R. H. H. 9/23/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26142

Analyst: rdh Instrument: BA000 Book # 11

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C251 SEG.#1 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000509	BY-105 (2)	3 SAMPLE	S98T002719	0	DLIQVOL1	SOLID	<u>1/A</u>	<u>250.0</u>		ml
98000509	BY-105 (2)	4 SAMPLE	S98T002719	0	DLIQWT01	SOLID	<u>1/A</u>	<u>357.0</u>		g
98000509	BY-105 (2)	5 SAMPLE	S98T002719	0	EST.G/ML	SOLID	<u>1/A</u>	<u>1.43</u>		g/ml
98000509	BY-105 (2)	6 SAMPLE	S98T002719	0	EXTRUD01	SOLID	<u>1/A</u>	<u>Complete</u>		
98000509	BY-105 (2)	7 SAMPLE	S98T002719	0	LLIQWT01	SOLID	<u>1/A</u>	<u>0</u>		g
98000509	BY-105 (2)	8 SAMPLE	S98T002719	0	NOTEBOOK	SOLID	<u>1/A</u>	<u>na</u>		
98000509	BY-105 (2)	9 SAMPLE	S98T002719	0	SLDWT-01	SOLID	<u>1/A</u>	<u>0</u>		g
98000509	BY-105 (2)	10 SAMPLE	S98T002719	0	APPEAR01	SOLID	<u>1/A</u>	<u>Complete</u>		
98000509	BY-105 (2)	11 SAMPLE	S98T002719	0	ORGVOL01	SOLID	<u>1/A</u>	<u>0</u>		ml

Final page for worklist # 26142

Ala Campbell 9/22/98
Analyst Signature Date

Ala Campbell 9/22/98
Analyst Signature Date
Approved RW Schaefer 9/23/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26145

Analyst: PSH Instrument: BA000 Hook # NA

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C251 SEG.#2 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000509	BY-105 (2)	3 SAMPLE	S98T002720	0	DLIQVOL1	SOLID	<u>1/A</u>	<u>250</u>		ml
98000509	BY-105 (2)	4 SAMPLE	S98T002720	0	DLIQWT01	SOLID	<u>1/A</u>	<u>334.7</u>		g
98000509	BY-105 (2)	5 SAMPLE	S98T002720	0	EST.G/ML	SOLID	<u>1/A</u>	<u>1.34</u>		g/ml
98000509	BY-105 (2)	6 SAMPLE	S98T002720	0	EXTRUD01	SOLID	<u>1/A</u>	<u>COMPLETE</u>		
98000509	BY-105 (2)	7 SAMPLE	S98T002720	0	LLIQWT01	SOLID	<u>N/A</u>	<u>LS</u>	<u>DID NOT RETAIN</u>	g
98000509	BY-105 (2)	8 SAMPLE	S98T002720	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>NA</u>		
98000509	BY-105 (2)	9 SAMPLE	S98T002720	0	SLDWT-01	SOLID	<u>N/A</u>	<u>LS</u>	<u>402.0</u>	<u>RET 9/22/98</u>
98000509	BY-105 (2)	10 SAMPLE	S98T002720	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000509	BY-105 (2)	11 SAMPLE	S98T002720	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 26145

Analyst Signature PSH Date 9/22/98

Analyst Signature PSH Date 9/22/98

Approved RW Schoeden 9/23/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26146

Analyst: R.D.H. Instrument: BA000 Look # NA

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C251 SEG#3 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.10</u>	<u>N/A</u>	
98000509	BY-105 (2)	3 SAMPLE	S98T002721	0	DLIQVOL1	SOLID	<u>N/A</u>	<u>250</u>		mL
98000509	BY-105 (2)	4 SAMPLE	S98T002721	0	DLIQWT01	SOLID	<u>N/A</u>	<u>365.1</u>		g
98000509	BY-105 (2)	5 SAMPLE	S98T002721	0	EST.G/ML	SOLID	<u>N/A</u>	<u>146</u>		g/mL
98000509	BY-105 (2)	6 SAMPLE	S98T002721	0	EXTRUD01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000509	BY-105 (2)	7 SAMPLE	S98T002721	0	LLIQWT01	SOLID	<u>N/A</u>	<u>40</u>	<u>DID NOT RETAIN</u>	g
98000509	BY-105 (2)	8 SAMPLE	S98T002721	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>NA</u>		
98000509	BY-105 (2)	9 SAMPLE	S98T002721	0	SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000509	BY-105 (2)	10 SAMPLE	S98T002721	0	APPEAR01	SOLID	<u>N/A</u>	<u>COMPLETE</u>		
98000509	BY-105 (2)	11 SAMPLE	S98T002721	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 26146

R.D.H. 9/22/98
Analyst Signature Date

R.D.H. 9/22/98
Analyst Signature Date
Approved R.D. Schwab 9/23/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26147

Analyst: drj Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C251 SEG.#4 RISER#7 EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>580.09</u>	<u>500.11</u>	<u>N/A</u>	
98000509	BY-105 (2)	3	SAMPLE	S98T002722	0		DLIQVOL1	SOLID	<u>1/A</u>	<u>250</u>		mL
98000509	BY-105 (2)	4	SAMPLE	S98T002722	0		DLIQWT01	SOLID	<u>1/A</u>	<u>330.8</u>		g
98000509	BY-105 (2)	5	SAMPLE	S98T002722	0		EST.G/ML	SOLID	<u>1/A</u>	<u>1.32</u>		g/mL
98000509	BY-105 (2)	6	SAMPLE	S98T002722	0		EXTRUD01	SOLID	<u>1/A</u>	<u>complete</u>		
98000509	BY-105 (2)	7	SAMPLE	S98T002722	0		LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000509	BY-105 (2)	8	SAMPLE	S98T002722	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>nr 995</u>		
98000509	BY-105 (2)	9	SAMPLE	S98T002722	0		SLDWT-01	SOLID	<u>N/A</u>	<u>103.6</u>		g
98000509	BY-105 (2)	10	SAMPLE	S98T002722	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000509	BY-105 (2)	11	SAMPLE	S98T002722	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 26147

D. R. Incha 9-27-98
Analyst Signature Date

D. R. Incha 9-27-98
Analyst Signature Date
Drj 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26148

Analyst: ABC Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod DO

Worklist Comment: BY-105 C251 SEG#4A RISER#7

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	<u>500.00</u>	<u>500.11</u>	N/A	
98000509	BY-105 (2)	3 SAMPLE	S98T002723	0	DLTQVOL1	SOLID	N/A	<u>0</u>		ml
98000509	BY-105 (2)	4 SAMPLE	S98T002723	0	DLTQWT01	SOLID	N/A	<u>0</u>		g
98000509	BY-105 (2)	5 SAMPLE	S98T002723	0	EST.G/ML	SOLID	N/A	<u>0</u>		g/ml
98000509	BY-105 (2)	6 SAMPLE	S98T002723	0	EXTRUD01	SOLID	N/A	<u>Complete</u>		
98000509	BY-105 (2)	7 SAMPLE	S98T002723	0	LLTQWT01	SOLID	N/A	<u>0</u>		g
98000509	BY-105 (2)	8 SAMPLE	S98T002723	0	NOTEBOOK	SOLID	N/A	<u>995.2</u>		
98000509	BY-105 (2)	9 SAMPLE	S98T002723	0	SLDWT-01	SOLID	N/A	<u>31.4</u>		g
98000509	BY-105 (2)	10 SAMPLE	S98T002723	0	APPEAR01	SOLID	N/A	<u>Complete</u>		
98000509	BY-105 (2)	11 SAMPLE	S98T002723	0	ORGVOL01	SOLID	N/A	<u>0</u>		ml

Final page for worklist # 26148

Alan B Campbell 9/27/98
Analyst Signature Date

Alan B Campbell 9/27/98
Analyst Signature Date
ABC 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26149

Analyst: ABC Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C251 SEG#5 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	<u>20.10</u>	<u>20.01</u>	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	<u>50.00</u>	<u>500.11</u>	N/A	
98000509	BY-105 (2)	3 SAMPLE	S98T002724	0	DLIQVOL1	SOLID	<u>1/A</u>	<u>10.0</u>		mL
98000509	BY-105 (2)	4 SAMPLE	S98T002724	0	DLIQWT01	SOLID	<u>1/A</u>	<u>128.8</u>		g
98000509	BY-105 (2)	5 SAMPLE	S98T002724	0	EST.G/ML	SOLID	<u>1/A</u>	<u>1.29</u>		g/mL
98000509	BY-105 (2)	6 SAMPLE	S98T002724	0	EXTRUD01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000509	BY-105 (2)	7 SAMPLE	S98T002724	0	LLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000509	BY-105 (2)	8 SAMPLE	S98T002724	0	NOTEBOOK	SOLID	<u>N/A</u>	<u>993.2</u>		
98000509	BY-105 (2)	9 SAMPLE	S98T002724	0	SLDWT-01	SOLID	<u>N/A</u>	<u>289.1</u>		g
98000509	BY-105 (2)	10 SAMPLE	S98T002724	0	APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000509	BY-105 (2)	11 SAMPLE	S98T002724	0	ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		mL

Final page for worklist # 26149

Alan B Campbell 9-27-98
Analyst Signature Date

Alan B Campbell 9/27/98
Analyst Signature Date
Aug 1 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26150

Analyst: ABC Instrument: BA000 Hook # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C252 SEG.#5 RISER#7 EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.11</u>	<u>N/A</u>	
98000510	BY-105 (2)	3	SAMPLE	S98T002731	0		DLIQVOL1	SOLID	<u>1/A</u>	<u>250</u>		ml
98000510	BY-105 (2)	4	SAMPLE	S98T002731	0		DLIQWT01	SOLID	<u>1/A</u>	<u>326.0</u>		g
98000510	BY-105 (2)	5	SAMPLE	S98T002731	0		EST.G/ML	SOLID	<u>1/A</u>	<u>130</u>		g/ml
98000510	BY-105 (2)	6	SAMPLE	S98T002731	0		EXTRUD01	SOLID	<u>1/A</u>	<u>Complete</u>		
98000510	BY-105 (2)	7	SAMPLE	S98T002731	0		LLIQWT01	SOLID	<u>1/A</u>	<u>0</u>		g
98000510	BY-105 (2)	8	SAMPLE	S98T002731	0		NOTEBOOK	SOLID	<u>1/A</u>	<u>995.2</u>		
98000510	BY-105 (2)	9	SAMPLE	S98T002731	0		SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000510	BY-105 (2)	10	SAMPLE	S98T002731	0		APPEAR01	SOLID	<u>N/A</u>	<u>Complete</u>		
98000510	BY-105 (2)	11	SAMPLE	S98T002731	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 26150

Alan Campbell 9/27/98
Analyst Signature Date

Alan Campbell 9/27/98
Analyst Signature Date
Ant 2 9/28/98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Lim., S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26151

Analyst: ABC Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY105 C252 SEG#6 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACT. AL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	20.00	20.01	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	500.00	500.11	N/A	
98000510	BY-105 (2)	3 SAMPLE	S98T002732	0	DLIQVOL1	SOLID	N/A	250		ml
98000510	BY-105 (2)	4 SAMPLE	S98T002732	0	DLIQWT01	SOLID	N/A	381.4		g
98000510	BY-105 (2)	5 SAMPLE	S98T002732	0	EST.G/ML	SOLID	N/A	1.53		g/ml
98000510	BY-105 (2)	6 SAMPLE	S98T002732	0	EXTRUD01	SOLID	N/A	Complete		
98000510	BY-105 (2)	7 SAMPLE	S98T002732	0	LLIQWT01	SOLID	N/A	0		g
98000510	BY-105 (2)	8 SAMPLE	S98T002732	0	NOTEBOOK	SOLID	N/A	995-2		
98000510	BY-105 (2)	9 SAMPLE	S98T002732	0	SLDWT-01	SOLID	N/A	0		g
98000510	BY-105 (2)	10 SAMPLE	S98T002732	0	APPEAR01	SOLID	N/A	Complete		
98000510	BY-105 (2)	11 SAMPLE	S98T002732	0	ORGVOL01	SOLID	N/A	0		ml

Final page for worklist # 26151

ABC Complete 9/27/98
Analyst Signature Date

ABC Complete 9/27/98
Analyst Signature Date
P. C. 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26152

Analyst: ABC Instrument: BA000 Hook # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY105 C252 SEG.#7 RISER#7 EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	20.00	20.01	N/A	
		2	INSTCHK02				EXTRUD01	SOLID	500.00	500.11	N/A	
98000510	BY-105 (2)	3	SAMPLE	S98T002733	0		DLIQVOL1	SOLID	N/A	250		ml
98000510	BY-105 (2)	4	SAMPLE	S98T002733	0		DLIQWT01	SOLID	N/A	361.4		g
98000510	BY-105 (2)	5	SAMPLE	S98T002733	0		EST.G/ML	SOLID	N/A	1.45		g/ml
98000510	BY-105 (2)	6	SAMPLE	S98T002733	0		EXTRUD01	SOLID	N/A	Complete		
98000510	BY-105 (2)	7	SAMPLE	S98T002733	0		LLIQWT01	SOLID	N/A	2		g
98000510	BY-105 (2)	8	SAMPLE	S98T002733	0		NOTEBOOK	SOLID	N/A	995.2		
98000510	BY-105 (2)	9	SAMPLE	S98T002733	0		SLDWT-01	SOLID	N/A	0		g
98000510	BY-105 (2)	10	SAMPLE	S98T002733	0		APPEAR01	SOLID	N/A	Complete		
98000510	BY-105 (2)	11	SAMPLE	S98T002733	0		ORGVOL01	SOLID	N/A	0		ml

Final page for worklist # 26152

ABC Campbell 9/27/98
Analyst Signature Date

ABC Campbell 9/27/98
Analyst Signature Date
Ante J. Garcia 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26153

Analyst: ABC Instrument: BA000 Book # _____

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C252 SEG#8 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	20.00	20.01	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	500.00	500.11	N/A	
98000510	BY-105 (2)	3 SAMPLE	S98T002734	0	DLIQVOL1	SOLID	1/A	250		ml
98000510	BY-105 (2)	4 SAMPLE	S98T002734	0	DLIQWT01	SOLID	1/A	386.2		g
98000510	BY-105 (2)	5 SAMPLE	S98T002734	0	EST.G/ML	SOLID	1/A	1.54		g/ml
98000510	BY-105 (2)	6 SAMPLE	S98T002734	0	EXTRUD01	SOLID	1/A	Complete		
98000510	BY-105 (2)	7 SAMPLE	S98T002734	0	LLIQWT01	SOLID	1/A	0		g
98000510	BY-105 (2)	8 SAMPLE	S98T002734	0	NOTEBOOK	SOLID	1/A	995.2		
98000510	BY-105 (2)	9 SAMPLE	S98T002734	0	SLDWT-01	SOLID	N/A	0		g
98000510	BY-105 (2)	10 SAMPLE	S98T002734	0	APPEAR01	SOLID	N/A	Complete		
98000510	BY-105 (2)	11 SAMPLE	S98T002734	0	ORGVOL01	SOLID	N/A	0		ml

Final page for worklist # 26153

ABC Campbell 9/27/98
Analyst Signature Date

ABC Campbell 9/27/98
Analyst Signature Date
ABC Campbell 9.28.98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

09/11/98 10:53

LABCORE Data Entry Template for Worklist# 26155 Analyst: ABC Instrument: BA000 Book # _____Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 C252 SEG.#9 RISER#7 EXTRUSION

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 INSTCHK01			EXTRUD01	SOLID	20.00	20.01	N/A	
		2 INSTCHK02			EXTRUD01	SOLID	500.00	500.11	N/A	
98000510	BY-105 (2)	3 SAMPLE	S98T012735	0	DLIQVOL1	SOLID	N/A	0		ml
98000510	BY-105 (2)	4 SAMPLE	S98T012735	0	DLIQWT01	SOLID	N/A	0		g
98000510	BY-105 (2)	5 SAMPLE	S98T012735	0	EST.G/ML	SOLID	N/A	0		g/ml
98000510	BY-105 (2)	6 SAMPLE	S98T012735	0	EXTRUD01	SOLID	N/A	Complete		
98000510	BY-105 (2)	7 SAMPLE	S98T012735	0	LLIQWT01	SOLID	N/A	0		g
98000510	BY-105 (2)	8 SAMPLE	S98T012735	0	NOTEBOOK	SOLID	N/A	995.2		
98000510	BY-105 (2)	9 SAMPLE	S98T012735	0	SLOWT-01	SOLID	N/A	0		g
98000510	BY-105 (2)	10 SAMPLE	S98T012735	0	APPEAR01	SOLID	N/A	Complete		
98000510	BY-105 (2)	11 SAMPLE	S98T012735	0	ORGVOL01	SOLID	N/A	0		ml

Final page for worklist # 26155

Alan B Campbell 9/27/98
 Analyst Signature Date

AB Campbell 9/27/98
 Analyst Signature Date
ABC 928.98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26217

Analyst: dci Instrument: BA000 Book # N/A

Method: LO-160-103 Rev/Mod D-0

Worklist Comment: BY-105 CORE 250 SEG. #8 RISER #7 EXTRUSION

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	INSTCHK01				EXTRUD01	SOLID	<u>20.00</u>	<u>20.01</u>	<u>N/A</u>	
		2	INSTCHK02				EXTRUD01	SOLID	<u>500.00</u>	<u>500.11</u>	<u>N/A</u>	
98000490	BY-105 (2)	3	SAMPLE	S98T002850	0		DLIQVOL1	SOLID	<u>N/A</u>	<u>0</u>		ml
98000490	BY-105 (2)	4	SAMPLE	S98T002850	0		DLIQWT01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	5	SAMPLE	S98T002850	0		EST.G/ML	SOLID	<u>N/A</u>	<u>0</u>		g/ml
98000490	BY-105 (2)	6	SAMPLE	S98T002850	0		EXTRUD01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	7	SAMPLE	S98T002850	0		LLIQWT01	SOLID	<u>N/A</u>	<u>15 did not spin</u>		
98000490	BY-105 (2)	8	SAMPLE	S98T002850	0		NOTEBOOK	SOLID	<u>N/A</u>	<u>WTC-N-995</u>		
98000490	BY-105 (2)	9	SAMPLE	S98T002850	0		SLDWT-01	SOLID	<u>N/A</u>	<u>0</u>		g
98000490	BY-105 (2)	10	SAMPLE	S98T002850	0		APPEAR01	SOLID	<u>N/A</u>	<u>complete</u>		
98000490	BY-105 (2)	11	SAMPLE	S98T002850	0		ORGVOL01	SOLID	<u>N/A</u>	<u>0</u>		ml

Final page for worklist # 26217

[Signature] 9-27-98
Analyst Signature Date

[Signature] 9-27-98
Analyst Signature Date
[Signature] 9-28-98

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

SAMPLE PREPARATION

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

LABCORE Data Entry Template for Worklist# 26816

Analyst: CRJ Instrument: FUS01 Boon # N/A

Method: LA-549-141 Rev/Mod G-0

Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>250</u>	<u>N/A</u>	<u>g/L</u>
98000436	BY-105 (2)	2 SAMPLE	S98T002598	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2.0372</u>		<u>g/L</u>
98000436	BY-105 (2)	3 SAMPLE	S98T002598	0	DOSE-02	SOLID	<u>N/A</u>	<u>2.05</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	4 DUP	S98T002598	0 F	FUSION01	SOLID	<u>2.0372</u>	<u>2.0374</u>	<u>N/A</u>	<u>g/L</u>
98000436	BY-105 (2)	5 DUP	S98T002598	0	DOSE-02	SOLID	<u>2.05</u>	<u>2.05</u>	<u>N/A</u>	<u>mrads/hour</u>
98000436	BY-105 (2)	6 SAMPLE	S98T002599	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2.0608</u>		<u>g/L</u>
98000436	BY-105 (2)	7 SAMPLE	S98T002599	0	DOSE-02	SOLID	<u>N/A</u>	<u>2.05</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	8 DUP	S98T002599	0 F	FUSION01	SOLID	<u>2.0608</u>	<u>2.0564</u>	<u>N/A</u>	<u>g/L</u>
98000436	BY-105 (2)	9 DUP	S98T002599	0	DOSE-02	SOLID	<u>2.05</u>	<u>2.05</u>	<u>N/A</u>	<u>mrads/hour</u>
98000436	BY-105 (2)	10 SAMPLE	S98T002600	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2.0164</u>		<u>g/L</u>
98000436	BY-105 (2)	11 SAMPLE	S98T002600	0	DOSE-02	SOLID	<u>N/A</u>	<u>2.05</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	12 DUP	S98T002600	0 F	FUSION01	SOLID	<u>2.0164</u>	<u>2.0084</u>	<u>N/A</u>	<u>g/L</u>
98000436	BY-105 (2)	13 DUP	S98T002600	0	DOSE-02	SOLID	<u>2.05</u>	<u>2.05</u>	<u>N/A</u>	<u>mrads/hour</u>

Final page for worklist # 26816

Analyst Signature

Date

Analyst Signature

Date

S98T002598 → S98T002598
2590 → 2599
2591 → 2500

Verified by: 8/1/98 10-29-98

Data Entry Comments:

HPT: Don CRAG 1000 10/2/98
added 30 mls of HNO₃

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26817

Analyst: gll Instrument: FUS01 Book # N/A

Method: LA-549-141 Rev/Mod B-0

Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>.250</u>	<u>N/A</u>	<u>g/L</u>
98000436	BY-105 (2)	2 SAMPLE	S98T002601	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2.0640</u>		<u>g/L</u>
			<u>0.5160 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	3 SAMPLE	S98T002601	0	DOSE-02	SOLID	<u>N/A</u>	<u>10</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	4 DUP	S98T002601	0 F	FUSION01	SOLID	<u>2.0640</u>	<u>2.0600</u>	<u>N/A</u>	<u>g/L</u>
			<u>0.5150 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	5 DUP	S98T002601	0	DOSE-02	SOLID	<u>10</u>	<u>10</u>	<u>N/A</u>	<u>mrads/hour</u>
98000436	BY-105 (2)	6 SAMPLE	S98T002602	0 F	FUSION01	SOLID	<u>N/A</u>	<u>2.0816</u>		<u>g/L</u>
			<u>0.5204 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	7 SAMPLE	S98T002602	0	DOSE-02	SOLID	<u>N/A</u>	<u>190</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	8 DUP	S98T002602	0 F	FUSION01	SOLID	<u>2.0816</u>	<u>2.0960</u>	<u>N/A</u>	<u>g/L</u>
			<u>0.5250 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	9 DUP	S98T002602	0	DOSE-02	SOLID	<u>190</u>	<u>190</u>	<u>N/A</u>	<u>mrads/hour</u>
98000436	BY-105 (2)	10 SAMPLE	S98T002603	0 F	FUSION01	SOLID	<u>N/A</u>	<u>1.9980</u>		<u>g/L</u>
			<u>0.4995 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	11 SAMPLE	S98T002603	0	DOSE-02	SOLID	<u>N/A</u>	<u>300</u>		<u>mrads/hour</u>
98000436	BY-105 (2)	12 DUP	S98T002603	0 F	FUSION01	SOLID	<u>1.9980</u>	<u>1.9648</u>	<u>N/A</u>	<u>g/L</u>
			<u>0.4912 g</u> →		<u>.250 L</u>					
98000436	BY-105 (2)	13 DUP	S98T002603	0	DOSE-02	SOLID	<u>300</u>	<u>300</u>	<u>N/A</u>	<u>mrads/hour</u>

Final page for worklist # 26817

Analyst Signature gll Date 10-27-98

Analyst Signature MLB Date 10/29/98

S98T002592 → S98T002601 Verified by: SLH 10-29-98
2593 → 2602
2594 → 2603

Data Entry Comments: HPT: Don CRAG 1000 10/29/98
ADDED 20 mL of NITRIC TO 2601
ADDED 40 mL of NITRIC TO 2602 + 2603

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26818

Analyst: gll Instrument: FUS01 Book # N/A

Method: LA-549-141 Rev/Mod G-D

Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	1	250	N/A	g/L
98000436	BY-105 (2)	2 SAMPLE	S98T002712 O F		FUSION01	SOLID	N/A	1.9668		g/L
			0.4917		→ .250					
98000436	BY-105 (2)	3 SAMPLE	S98T002712 O		DOSE-02	SOLID	N/A	37.5		mrad/hour
98000436	BY-105 (2)	4 DUP	S98T002712 O F		FUSION01	SOLID	1.9668	1.9616	N/A	g/L
			0.4904		→ .250					
98000436	BY-105 (2)	5 DUP	S98T002712 O		DOSE-02	SOLID	37.5	37.5	N/A	mrad/hour
98000490	BY-105 (2)	6 SAMPLE	S98T002739 O F		FUSION01	SOLID	N/A	2.0048		g/L
			0.5012		→ .250					
98000490	BY-105 (2)	7 SAMPLE	S98T002739 O		DOSE-02	SOLID	N/A	1.5		mrad/hour
98000490	BY-105 (2)	8 DUP	S98T002739 O F		FUSION01	SOLID	2.0048	2.0124	N/A	g/L
			0.5031		→ .250					
98000490	BY-105 (2)	9 DUP	S98T002739 O		DOSE-02	SOLID	1.5	1.5	N/A	mrad/hour
98000437	BY-105 (2)	10 SAMPLE	S98T002844 O F		FUSION01	SOLID	N/A	1.9852		g/L
			0.4963		→ .250					
98000437	BY-105 (2)	11. SAMPLE	S98T002844 O		DOSE-02	SOLID	N/A	21		mrad/hour
98000437	BY-105 (2)	12 DUP	S98T002844 O F		FUSION01	SOLID	1.9852	1.9944	N/A	g/L
			1.4986		→ .250					
98000437	BY-105 (2)	13 DUP	S98T002844 O		DOSE-02	SOLID	21	21	N/A	mrad/hour

Final page for worklist # 26818

Analyst Signature gll Date 11-4-98

Analyst Signature Julia Z. Chao Date 11-9-98

7 598T002711 → 598T002712 Verified by: MB 11/9/98
10 2737 → 2739
1 2840 → 2844

Data Entry Comments: Hpt: Sam Salazar
added 20 mls of Nitric

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26819

Analyst: gll Instrument: FUS01 Book: # NA
Method: LA-549-141 Rev/Mod G-0
Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>.250</u>	<u>N/A</u>	<u>g/L</u>
98000437	BY-105 (2)	2 SAMPLE	S98T002845 0 F <u>0.5095</u>		FUSION01	SOLID	<u>N/A</u>	<u>2.0380</u>		<u>g/L</u>
98000437	BY-105 (2)	3 SAMPLE	S98T002845 0		DOSE-02	SOLID	<u>N/A</u>	<u>86.5</u>		<u>mrad/hour</u>
98000437	BY-105 (2)	4 DUP	S98T002845 0 F <u>0.5038</u>		FUSION01	SOLID	<u>2.0380</u>	<u>2.0152</u>	<u>N/A</u>	<u>g/L</u>
98000437	BY-105 (2)	5 DUP	S98T002845 0		DOSE-02	SOLID	<u>86.5</u>	<u>110</u>	<u>N/A</u>	<u>mrad/hour</u>
98000490	BY-105 (2)	6 SAMPLE	S98T002918 0 F <u>0.5021</u>		FUSION01	SOLID	<u>N/A</u>	<u>2.0084</u>		<u>g/L</u>
98000490	BY-105 (2)	7 SAMPLE	S98T002918 0		DOSE-02	SOLID	<u>N/A</u>	<u>2.4</u>		<u>mrad/hour</u>
98000490	BY-105 (2)	8 DUP	S98T002918 0 F <u>0.5096</u>		FUSION01	SOLID	<u>2.0084</u>	<u>2.0384</u>	<u>N/A</u>	<u>g/L</u>
98000490	BY-105 (2)	9 DUP	S98T002918 0		DOSE-02	SOLID	<u>2.4</u>	<u>1.5</u>	<u>N/A</u>	<u>mrad/hour</u>
98000490	BY-105 (2)	10 SAMPLE	S98T002919 0 F <u>0.4990</u>		FUSION01	SOLID	<u>N/A</u>	<u>1.9960</u>		<u>g/L</u>
98000490	BY-105 (2)	11 SAMPLE	S98T002919 0		DOSE-02	SOLID	<u>N/A</u>	<u>20.5</u>		<u>mrad/hour</u>
98000490	BY-105 (2)	12 DUP	S98T002919 0 F <u>0.4925</u>		FUSION01	SOLID	<u>1.9960</u>	<u>1.9700</u>	<u>N/A</u>	<u>g/L</u>
98000490	BY-105 (2)	13 DUP	S98T002919 0		DOSE-02	SOLID	<u>20.5</u>	<u>20.5</u>	<u>N/A</u>	<u>mrad/hour</u>

Final page for worklist # 26819

gll 10-30-98
Analyst Signature Date

gll 11-2-98
Analyst Signature Date

598T002841 → 598T002845
2910 → 2918
2911 → 2919

Data Entry Comments:

HPT: Kathy Alessi
Added 30 nls of Nitric in 2845 Sam + Dup - 20 nls in 2918
ad 2919 Sam + Dup.

Verified by: gll 11/3/98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26820

Analyst: gll Instrument: FUS01 GO Book # NA
Method: LA-549-141 Rev/Mod G-0
Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP		FUSION01	SOLID	<u>1</u>	<u>250</u>	N/A	g/L
98000490	BY-105 (2)	2	SAMPLE	S98T002920	0 F <u>0.5104g</u> → <u>250L</u>	FUSION01	SOLID	<u>N/A</u>	<u>2.0416</u>	g/L
98000490	BY-105 (2)	3	SAMPLE	S98T002920	0 <u>DOSE-02</u>	SOLID	<u>N/A</u>	<u>2.05</u>		mrad/hour
98000490	BY-105 (2)	4	DUP	S98T002920	0 F <u>0.5102g</u> → <u>250L</u>	FUSION01	SOLID	<u>2.0411</u>	<u>2.0408</u>	N/A
98000490	BY-105 (2)	5	DUP	S98T002920	0 <u>DOSE-02</u>	SOLID	<u>2.05</u>	<u>2.05</u>	N/A	mrad/hour
98000490	BY-105 (2)	6	SAMPLE	S98T002921	0 F <u>0.5095</u> → <u>250L</u>	FUSION01	SOLID	<u>N/A</u>	<u>2.0380</u>	g/L
98000490	BY-105 (2)	7	SAMPLE	S98T002921	0 <u>DOSE-02</u>	SOLID	<u>N/A</u>	<u>1.6</u>		mrad/hour
98000490	BY-105 (2)	8	DUP	S98T002921	0 F <u>0.5060</u> → <u>250L</u>	FUSION01	SOLID	<u>2.0380</u>	<u>2.0240</u>	N/A
98000490	BY-105 (2)	9	DUP	S98T002921	0 <u>DOSE-02</u>	SOLID	<u>1.6</u>	<u>1.6</u>	N/A	mrad/hour
98000490	BY-105 (2)	10	SAMPLE	S98T002922	0 F <u>0.4902</u> → <u>250L</u>	FUSION01	SOLID	<u>N/A</u>	<u>1.9608</u>	g/L
98000490	BY-105 (2)	11	SAMPLE	S98T002922	0 <u>DOSE-02</u>	SOLID	<u>N/A</u>	<u>91</u>		mrad/hour
98000490	BY-105 (2)	12	DUP	S98T002922	0 F <u>0.4917</u> → <u>250L</u>	FUSION01	SOLID	<u>1.9608</u>	<u>1.9668</u>	N/A
98000490	BY-105 (2)	13	DUP	S98T002922	0 <u>DOSE-02</u>	SOLID	<u>91</u>	<u>58</u>	N/A	mrad/hour

Final page for worklist # 26820

Analyst Signature [Signature] Date 10-29-98

Analyst Signature [Signature] Date 11/10/98

5987002912 → 5987002920
2913 → 2921
2914 → 2922
Verified by:

Data Entry Comments:

Added 20 ml of Nitric in 2920 & 2921 - 30ml in 2922
Hpt: Kathy Alsin

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

faxed to 373-1180 @ 1445 10/29/98
faxed again to 373-1180 @ 1530 11-9-98
298
[Signature]

LABCORE Data Entry Template for Worklist# 26821

Analyst: gll Instrument: FUS01 Book # NA
Method: LA-549-141 Rev/Mod GND
Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLER#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>1.250</u>	<u>N/A</u>	<u>g/L</u>
98000509	BY-105 (2)	2 SAMPLE	S98T002986 0 F <u>0.4930</u>		FUSION01 → <u>.250</u>	SOLID	<u>N/A</u>	<u>1.9720</u>		<u>g/L</u>
98000509	BY-105 (2)	3 SAMPLE	S98T002986 0		DOSE-02	SOLID	<u>N/A</u>	<u>6.75</u>		<u>mrad/hour</u>
98000509	BY-105 (2)	4 DUP	S98T002986 0 F <u>0.4960</u>		FUSION01 → <u>.250</u>	SOLID	<u>1.9720</u>	<u>1.9840</u>	<u>N/A</u>	<u>g/L</u>
98000509	BY-105 (2)	5 DUP	S98T002986 0		DOSE-02	SOLID	<u>6.75</u>	<u>6.75</u>	<u>N/A</u>	<u>mrad/hour</u>
98000509	BY-105 (2)	6 SAMPLE	S98T002987 0 F <u>0.5068</u>		FUSION01 → <u>.250</u>	SOLID	<u>N/A</u>	<u>2.0272</u>		<u>g/L</u>
98000509	BY-105 (2)	7 SAMPLE	S98T002987 0		DOSE-02	SOLID	<u>N/A</u>	<u>7.5</u>		<u>mrad/hour</u>
98000509	BY-105 (2)	8 DUP	S98T002987 0 F <u>0.5078</u>		FUSION01 → <u>.250</u>	SOLID	<u>2.0272</u>	<u>2.0312</u>	<u>N/A</u>	<u>g/L</u>
98000509	BY-105 (2)	9 DUP	S98T002987 0		DOSE-02	SOLID	<u>7.5</u>	<u>7.5</u>	<u>N/A</u>	<u>mrad/hour</u>

Final page for worklist # 26821

Analyst Signature S. L. L. L. T. Date 11-4-98

Analyst Signature Julia Chan Date 11/5/98

Verified by ML 11/5/98

S98T002980 → S98T002986
2981 → 2987
2982 →

Data Entry Comments:

Hof: Sam Salazar
added 20 rls of Netire.

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26822

Analyst: gll Instrument: FUS01 Book # NA

Method: LA-549-141 Rev/Mod G-B

Worklist Comment: BY-105 FOR FUSION01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP		FUSION01	SOLID	<u>1</u>	<u>1.250</u>	<u>N/A</u>	<u>g/L</u>
98000509	BY-105 (2)	2	SAMPLE	S98T002988	O F <u>0.4968</u> → <u>0.250</u>	FUSION01	SOLID	<u>N/A</u>	<u>1.9872</u>	<u>g/L</u>
98000509	BY-105 (2)	3	SAMPLE	S98T002988	O DOSE-02	SOLID	<u>N/A</u>	<u>0.5</u>		<u>mrad/hr</u>
98000509	BY-105 (2)	4	DUP	S98T002988	O F <u>0.4998</u> → <u>0.250</u>	FUSION01	SOLID	<u>1.9872</u>	<u>1.9992</u>	<u>g/L</u>
98000509	BY-105 (2)	5	DUP	S98T002988	O DOSE-02	SOLID	<u>0.5</u>	<u>2.6</u>	<u>N/A</u>	<u>mrad/hr</u>
98000509	BY-105 (2)	6	SAMPLE	S98T002989	O F <u>0.5048</u> → <u>0.250</u>	FUSION01	SOLID	<u>N/A</u>	<u>2.0192</u>	<u>g/L</u>
98000509	BY-105 (2)	7	SAMPLE	S98T002989	O DOSE-02	SOLID	<u>N/A</u>	<u>4.5</u>		<u>mrad/hr</u>
98000509	BY-105 (2)	8	DUP	S98T002989	O F <u>0.5057</u> → <u>0.250</u>	FUSION01	SOLID	<u>2.0192</u>	<u>2.0228</u>	<u>g/L</u>
98000509	BY-105 (2)	9	DUP	S98T002989	O DOSE-02	SOLID	<u>4.5</u>	<u>4.5</u>	<u>N/A</u>	<u>mrad/hr</u>

Final page for worklist # 26822

Analyst Signature [Signature] Date 11/5/98

Analyst Signature [Signature] Date 11-9-98

598700 2982 → 598700 2988
2983 → 2989

Data Entry Comments:

HPT: Kathy Alarcon
added 20 mls of Nitric

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27651

Analyst: gll Instrument: FUS01 Book # NA

Method: LA-549-141 Rev/Mod 2-0

Worklist Comment: BY-105 FUSION01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			FUSION01	SOLID	<u>1</u>	<u>.250</u>	N/A	g/L
98000490	BY-105 (2)	2 SAMPLE	S98T003388 0 F		FUSION01	SOLID	N/A	<u>2.0220</u>		g/L
98000490	BY-105 (2)	3 SAMPLE	S98T003388 0	<u>0.5855</u>	DOSE-02	SOLID	<u>.250</u>	N/A	<u>1.5</u>	mrad/hour
98000490	BY-105 (2)	4 DUP	S98T003388 0 F		FUSION01	SOLID	<u>2.0220</u>	<u>2.0240</u>	N/A	g/L
98000490	BY-105 (2)	5 DUP	S98T003388 0	<u>0.5860</u>	DOSE-02	SOLID	<u>.250</u>	<u>1.5</u>	<u>1.5</u>	mrad/hour
98000436	BY-105 (2)	6 SAMPLE	S98T003393 0 F		FUSION01	SOLID	N/A	<u>2.0140</u>		g/L
98000436	BY-105 (2)	7 SAMPLE	S98T003393 0	<u>0.5035</u>	DOSE-02	SOLID	<u>.250</u>	N/A	<u><0.5</u>	mrad/hour
98000436	BY-105 (2)	8 DUP	S98T003393 0 F		FUSION01	SOLID	<u>2.0140</u>	<u>2.0312</u>	<u>2.0312</u>	g/L
98000436	BY-105 (2)	9 DUP	S98T003393 0	<u>0.5078</u>	DOSE-02	SOLID	<u>.250</u>	<u><0.5</u>	<u><0.5</u>	mrad/hour
98000436	BY-105 (2)	10 SAMPLE	S98T003398 0 F		FUSION01	SOLID	N/A	<u>1.9860</u>		g/L
98000436	BY-105 (2)	11 SAMPLE	S98T003398 0	<u>0.4965</u>	DOSE-02	SOLID	<u>.250</u>	N/A	<u>200</u>	mrad/hour
98000436	BY-105 (2)	12 DUP	S98T003398 0 F		FUSION01	SOLID	<u>1.9860</u>	<u>1.9840</u>	N/A	g/L
98000436	BY-105 (2)	13 DUP	S98T003398 0	<u>0.4960</u>	DOSE-02	SOLID	<u>.250</u>	<u>200</u>	<u>200</u>	mrad/hour

Final page for worklist # 27651

Analyst Signature gll Date 1-7-99

Verified by: MLB Analyst Signature MLB Date 1/8/99

598700 3386 → 598700 3388
598700 3416 → 598700 3393
598700 3415 → 598700 3398

Data Entry Comments: HIT: Jim Hatcher

added 20 uLs of Nitric in 3388 + 3393
added 30 uLs of Nitric in 3398

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26824

Analyst: ADP Instrument: H2001 Book #   Method: LA-504-101 Rev/Mod 60

Worklist Comment: BY-105 FOR H20DIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H20DIG01	SOLID	<u>1</u>	<u>100</u>	N/A	g/L
98000436	BY-105 (2)	2 SAMPLE	A ¹² S98T002610 0 W		H20DIG01	SOLID	N/A	<u>5.304</u>		g/L
98000436	BY-105 (2)	3 SAMPLE	<u>5304</u> <u>.5328</u> → <u>.100L</u> S98T002610 0		DOSE-02	SOLID	N/A	<u>1.5</u>		mrad/hour
98000436	BY-105 (2)	4 DUP	S98T002610 0 W		H20DIG01	SOLID	<u>5304</u>	<u>5328</u>	N/A	g/L
98000436	BY-105 (2)	5 DUP	<u>.5328</u> → <u>.100L</u> S98T002610 0		DOSE-02	SOLID	<u>1.5</u>	<u>1.5</u>	N/A	mrad/hour
98000436	BY-105 (2)	6 SAMPLE	S98T002611 0 W		H20DIG01	SOLID	N/A	<u>5.218</u>		g/L
98000436	BY-105 (2)	7 SAMPLE	<u>.5218</u> → <u>.100L</u> S98T002611 0		DOSE-02	SOLID	N/A	<u>1.5</u>		mrad/hour
98000436	BY-105 (2)	8 DUP	S98T002611 0 W		H20DIG01	SOLID	<u>5218</u>	<u>5210</u>	N/A	g/L
98000436	BY-105 (2)	9 DUP	<u>.5210</u> → <u>.100L</u> S98T002611 0		DOSE-02	SOLID	<u>1.5</u>	<u>1.5</u>	N/A	mrad/hour
98000436	BY-105 (2)	10 SAMPLE	S98T002612 0 W		H20DIG01	SOLID	N/A	<u>5.241</u>		g/L
98000436	BY-105 (2)	11 SAMPLE	<u>.5241</u> → <u>.100L</u> S98T002612 0		DOSE-02	SOLID	N/A	<u>.5</u>		mrad/hour
98000436	BY-105 (2)	12 DUP	S98T002612 0 W		H20DIG01	SOLID	<u>5241</u>	<u>5228</u>	N/A	g/L
98000436	BY-105 (2)	13 DUP	<u>.5228</u> → <u>.100L</u> S98T002612 0		DOSE-02	SOLID	<u>.5</u>	<u>.5</u>	N/A	mrad/hour

Final page for worklist # 26824

Analyst Signature Anthony Punton Date 12-1-98

Analyst Signature _____ Date _____

S98T002589 → 2610
 2590 → 2611
 2591 → 2612

Data Entry Comments:

DATA ENTRY 12-2-98

John Howell

Validated by
J. Ananth 12-10-98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

worklistrpt Version 2.1 05/15/95
10/26/98 10:47

Page: 1

LABCORE Data Entry Template for Worklist# 26826

Analyst: Pat Instrument: H2001 Book # 114Method: LA-504-101 Rev/Mod 6-6

Worklist Comment: BY-105 FOR H2ODIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2001G01	SOLID	<u>1</u>	<u>.108</u>	N/A	g/L
98000436	BY-105 (2)	2 SAMPLE	S98T002613	0 W	H2001G01	SOLID	N/A	<u>4.945</u>		g/L
98000436	BY-105 (2)	3 SAMPLE	S98T002613	0	DOSE-02	SOLID	N/A	<u>2.5</u>		mrad/hour
98000436	BY-105 (2)	4 DUP	S98T002613	0 W	H2001G01	SOLID	<u>4.945</u>	<u>5.041</u>	N/A	g/L
98000436	BY-105 (2)	5 DUP	S98T002613	0	DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	N/A	mrad/hour
98000436	BY-105 (2)	6 SAMPLE	S98T002614	0 W	H2001G01	SOLID	N/A	<u>5.124</u>		g/L
98000436	BY-105 (2)	7 SAMPLE	S98T002614	0	DOSE-02	SOLID	N/A	<u>6</u>		mrad/hour
98000436	BY-105 (2)	8 DUP	S98T002614	0 W	H2001G01	SOLID	<u>5.124</u>	<u>4.968</u>	N/A	g/L
98000436	BY-105 (2)	9 DUP	S98T002614	0	DOSE-02	SOLID	<u>6</u>	<u>6</u>	N/A	mrad/hour
98000436	BY-105 (2)	10 SAMPLE	S98T002615	0 W	H2001G01	SOLID	N/A	<u>4.995</u>		g/L
98000436	BY-105 (2)	11 SAMPLE	S98T002615	0	DOSE-02	SOLID	N/A	<u>8</u>		mrad/hour
98000436	BY-105 (2)	12 DUP	S98T002615	0 W	H2001G01	SOLID	<u>4.995</u>	<u>5.149</u>	N/A	g/L
98000436	BY-105 (2)	13 DUP	S98T002615	0	DOSE-02	SOLID	<u>8</u>	<u>8</u>	N/A	mrad/hour

Final page for worklist # 26826

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

DATA ENTRY 12-31-98

John W. Powell

Validated by:

Sam H. Perry

01/07/99

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26828

Analyst: Jan 2 Instrument: H2001 Book # N/A

Method: LA-504-101 Rev/Mod G-6

Worklist Comment: BY-105 FOR H2ODIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2ODIG01	SOLID	<u>1</u>	<u>.100</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	2 SAMPLE	S98T002714	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.099</u>		g/L
		<u>5099.7</u> → <u>.100L</u>								
98000436	BY-105 (2)	3 SAMPLE	S98T002714	0	DOSE-02	SOLID	<u>N/A</u>	<u>6</u>		mrad/hour
98000436	BY-105 (2)	4 DUP	S98T002714	0 W	H2ODIG01	SOLID	<u>5.099</u>	<u>5.167</u>	<u>N/A</u>	g/L
		<u>5167.7</u> → <u>.100L</u>								
98000436	BY-105 (2)	5 DUP	S98T002714	0	DOSE-02	SOLID	<u>6</u>	<u>6</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	6 SAMPLE	S98T002741	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>4.993</u>		g/L
		<u>4993.7</u> → <u>.100L</u>								
98000490	BY-105 (2)	7 SAMPLE	S98T002741	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000490	BY-105 (2)	8 DUP	S98T002741	0 W	H2ODIG01	SOLID	<u>4.993</u>	<u>4.951</u>	<u>N/A</u>	g/L
		<u>4951.5</u> → <u>.100L</u>								
98000490	BY-105 (2)	9 DUP	S98T002741	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	mrad/hour
98000437	BY-105 (2)	10 SAMPLE	S98T002848	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.108</u>		g/L
		<u>5108.7</u> → <u>.100L</u>								
98000437	BY-105 (2)	11 SAMPLE	S98T002848	0	DOSE-02	SOLID	<u>N/A</u>	<u>3</u>		mrad/hour
98000437	BY-105 (2)	12 DUP	S98T002848	0 W	H2ODIG01	SOLID	<u>5.108</u>	<u>5.108</u>	<u>N/A</u>	g/L
		<u>5108.7</u> → <u>.100L</u>								
98000437	BY-105 (2)	13 DUP	S98T002848	0	DOSE-02	SOLID	<u>3</u>	<u>3</u>	<u>N/A</u>	mrad/hour

Final page for worklist # 26828

Analyst Signature [Signature] Date 12/23/98

Analyst Signature _____ Date _____

Data Entry Comments:

DATA ENTRY 12-20-98

[Signature]

Validated 12/28/98

[Signature]

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26830

Analyst: Jan 2 Instrument: H2001 Book # N/A
Method: LA-504-101 Rev/Mod G-0

Worklist Comment: BY-105 FOR H2ODIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2ODIG01	SOLID	<u>1</u>	<u>.100</u>	<u>N/A</u>	g/L
98000437	BY-105 (2)	2 SAMPLE	S98T002849	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>4.981</u>		g/L
		<u>.4981</u>	<u>→</u>	<u>.100</u>						
98000437	BY-105 (2)	3 SAMPLE	S98T002849	0	DOSE-02	SOLID	<u>N/A</u>	<u>3</u>		mrad/hour
98000437	BY-105 (2)	4 DUP	S98T002849	0 W	H2ODIG01	SOLID	<u>4.981</u>	<u>5.042</u>	<u>N/A</u>	g/L
		<u>.5042</u>	<u>→</u>	<u>.100</u>						
98000437	BY-105 (2)	5 DUP	S98T002849	0	DOSE-02	SOLID	<u>3</u>	<u>3</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	6 SAMPLE	S98T002931	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.031</u>		g/L
		<u>.5031</u>	<u>→</u>	<u>.100</u>						
98000490	BY-105 (2)	7 SAMPLE	S98T002931	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000490	BY-105 (2)	8 DUP	S98T002931	0 W	H2ODIG01	SOLID	<u>5.031</u>	<u>5.142</u>	<u>N/A</u>	g/L
		<u>.5142</u>	<u>→</u>	<u>.100</u>						
98000490	BY-105 (2)	9 DUP	S98T002931	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	10 SAMPLE	S98T002932	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.004</u>		g/L
		<u>.5004</u>	<u>→</u>	<u>.100</u>						
98000490	BY-105 (2)	11 SAMPLE	S98T002932	0	DOSE-02	SOLID	<u>N/A</u>	<u>4.5</u>		mrad/hour
98000490	BY-105 (2)	12 DUP	S98T002932	0 W	H2ODIG01	SOLID	<u>5.004</u>	<u>4.935</u>	<u>N/A</u>	g/L
		<u>.4935</u>	<u>→</u>	<u>.100</u>						
98000490	BY-105 (2)	13 DUP	S98T002932	0	DOSE-02	SOLID	<u>4.5</u>	<u>4.5</u>	<u>N/A</u>	mrad/hour

Final page for worklist # 26830

Jan 2 12/29/98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

DATA ENTRY 12-29-98
Jan 2

Validated 12/30/98 Jan 2

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26832

Analyst: P2 Instrument: H2001 Book # N/A

Method: LA-504-101 Rev/Mod G-6

Worklist Comment: BY-105 FOR H20DIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H20DIG01	SOLID	<u>1</u>	<u>100</u>	<u>N/A</u>	g/L
98000490	BY-105 (2)	2 SAMPLE	S98T002933	0 W	H20DIG01	SOLID	<u>N/A</u>	<u>5.047</u>		g/L
98000490	BY-105 (2)	3 SAMPLE	S98T002933	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000490	BY-105 (2)	4 DUP	S98T002933	0 W	H20DIG01	SOLID	<u>5.047</u>	<u>5.033</u>	<u>N/A</u>	g/L
98000490	BY-105 (2)	5 DUP	S98T002933	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	6 SAMPLE	S98T002934	0 W	H20DIG01	SOLID	<u>N/A</u>	<u>5.118</u>		g/L
98000490	BY-105 (2)	7 SAMPLE	S98T002934	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000490	BY-105 (2)	8 DUP	S98T002934	0 W	H20DIG01	SOLID	<u>5.118</u>	<u>5.086</u>	<u>N/A</u>	g/L
98000490	BY-105 (2)	9 DUP	S98T002934	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	10 SAMPLE	S98T002935	0 W	H20DIG01	SOLID	<u>N/A</u>	<u>4.966</u>		g/L
98000490	BY-105 (2)	11 SAMPLE	S98T002935	0	DOSE-02	SOLID	<u>N/A</u>	<u>2.5</u>		mrad/hour
98000490	BY-105 (2)	12 DUP	S98T002935	0 W	H20DIG01	SOLID	<u>4.966</u>	<u>4.974</u>	<u>N/A</u>	g/L
98000490	BY-105 (2)	13 DUP	S98T002935	0	DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	<u>N/A</u>	mrad/hour

Final page for worklist # 26832

Analyst Signature P. Collob Date 12/28/98

Analyst Signature _____ Date _____

Data Entry Comments:

DATA ENTRY 12-28-98
John Warrall

Validated 12/30/98 JML/eye

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

worklistrpt Version 2.1 05/15/95
10/26/98 10:51

Page: 1

LABCORE Data Entry Template for Worklist# 26834

Analyst: PA12 Instrument: H2001 Bottle # N/AMethod: LA-504-101 Rev/Mod G-0

Worklist Comment: BY-105 FOR H2ODIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			H2ODIG01	SOLID	<u>1</u>	<u>.100</u>	N/A	g/L
98000509	BY-105 (2)	2 SAMPLE	S98T002994	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.121</u>		g/L
98000509	BY-105 (2)	3 SAMPLE	S98T002994	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000509	BY-105 (2)	4 DUP	S98T002994	0 W	H2ODIG01	SOLID	<u>5.121</u>	<u>5.091</u>	N/A	g/L
98000509	BY-105 (2)	5 DUP	S98T002994	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	N/A	mrad/hour
98000509	BY-105 (2)	6 SAMPLE	S98T002995	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.072</u>		g/L
98000509	BY-105 (2)	7 SAMPLE	S98T002995	0	DOSE-02	SOLID	<u>N/A</u>	<u>1</u>		mrad/hour
98000509	BY-105 (2)	8 DUP	S98T002995	0 W	H2ODIG01	SOLID	<u>5.072</u>	<u>5.054</u>	N/A	g/L
98000509	BY-105 (2)	9 DUP	S98T002995	0	DOSE-02	SOLID	<u>1</u>	<u>1</u>	N/A	mrad/hour
98000509	BY-105 (2)	10 SAMPLE	S98T002996	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.051</u>		g/L
98000509	BY-105 (2)	11 SAMPLE	S98T002996	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000509	BY-105 (2)	12 DUP	S98T002996	0 W	H2ODIG01	SOLID	<u>5.051</u>	<u>5.098</u>	N/A	g/L
98000509	BY-105 (2)	13 DUP	S98T002996	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	N/A	mrad/hour
98000509	BY-105 (2)	14 SAMPLE	S98T002997	0 W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.080</u>		g/L
98000509	BY-105 (2)	15 SAMPLE	S98T002997	0	DOSE-02	SOLID	<u>N/A</u>	<u>1</u>		mrad/hour
98000509	BY-105 (2)	16 DUP	S98T002997	0 W	H2ODIG01	SOLID	<u>5.080</u>	<u>5.058</u>	N/A	g/L
98000509	BY-105 (2)	17 DUP	S98T002997	0	DOSE-02	SOLID	<u>1</u>	<u>1</u>	N/A	mrad/hour

Data Entry Comments:

HPT. Lisa Valdez

Validated by:

DATA ENTRY 11-19-98

S. M. Pang
11/20/98Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26834

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26834

11/19/98
_____
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27653

Analyst: Law Instrument: H2001 Book # W/AMethod: LA-504-101 Rev/Mod G-0

Worklist Comment: BY-105 H2ODIG01 RLC

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP				H2ODIG01	SOLID	<u>1</u>	<u>.100</u>	<u>N/A</u>	g/L
98000490	BY-105 (2)	2	SAMPLE	S98T003390	0	W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.036</u>		g/L
				<u>.5036 g</u>								
98000490	BY-105 (2)	3	SAMPLE	S98T003390	0		DOSE-02	SOLID	<u>N/A</u>	<u>.6</u>		mrad/hour
98000490	BY-105 (2)	4	DUP	S98T003390	0	W	H2ODIG01	SOLID	<u>5.036</u>	<u>5.133</u>	<u>N/A</u>	g/L
				<u>.5133 g</u>								
98000490	BY-105 (2)	5	DUP	S98T003390	0		DOSE-02	SOLID	<u>.6</u>	<u>.6</u>	<u>N/A</u>	mrad/hour
98000436	BY-105 (2)	6	SAMPLE	S98T003395	0	W	H2ODIG01	SOLID	<u>N/A</u>	<u>4.942</u>		g/L
				<u>.4742 g</u>								
98000436	BY-105 (2)	7	SAMPLE	S98T003395	0		DOSE-02	SOLID	<u>N/A</u>	<u>22.5</u>		mrad/hour
98000436	BY-105 (2)	8	DUP	S98T003395	0	W	H2ODIG01	SOLID	<u>4.942</u>	<u>5.158</u>	<u>N/A</u>	g/L
				<u>.5158 g</u>								
98000436	BY-105 (2)	9	DUP	S98T003395	0		DOSE-02	SOLID	<u>22.5</u>	<u>22.5</u>	<u>N/A</u>	mrad/hour
98000436	BY-105 (2)	10	SAMPLE	S98T003400	0	W	H2ODIG01	SOLID	<u>N/A</u>	<u>5.068</u>		g/L
				<u>.5068 g</u>								
98000436	BY-105 (2)	11	SAMPLE	S98T003400	0		DOSE-02	SOLID	<u>N/A</u>	<u>3</u>		mrad/hour
98000436	BY-105 (2)	12	DUP	S98T003400	0	W	H2ODIG01	SOLID	<u>5.068</u>	<u>5.072</u>	<u>N/A</u>	g/L
				<u>.5072 g</u>								
98000436	BY-105 (2)	13	DUP	S98T003400	0		DOSE-02	SOLID	<u>3</u>	<u>3</u>	<u>N/A</u>	mrad/hour

Final page for worklist # 27653

Analyst Signature Law Date 12/22/98

Analyst Signature _____ Date _____

S98T003386 → S98T003390

3391 → 3395

3396 → 3400

Data Entry Comments:

DATA ENTRY 12-23-98

John W. Wall

Validated 12/28/98

JB

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26823

Analyst: ADD Instrument: ACD01 Book # WAC-1A WAC-2 2.5mls EACH TO 50mL Final Volume DF=20X
Method: LA-505-163 Rev/Mod B-0
Worklist Comment: BY-105 FOR ACIDIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>1</u>	<u>0.050</u>	<u>N/A</u>	g/L
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>20</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	3 SAMPLE	S98T002604 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>5.428</u>		g/L
98000436	BY-105 (2)	4 SAMPLE	S98T002604 0		DOSE-02	SOLID	<u>N/A</u>	<u>1.5</u>		mrads/hour
98000436	BY-105 (2)	5 DUP	S98T002604 0 A		ACIDIG01	SOLID	<u>5.428</u>	<u>5.6</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	6 DUP	S98T002604 0		DOSE-02	SOLID	<u>1.5</u>	<u>1.5</u>	<u>N/A</u>	mrads/hour
98000436	BY-105 (2)	7 SPK	S98T002604 0 A		ACIDIG01	SOLID	<u>1.5</u>	<u>5.37</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	8 SPK	S98T002604 0		DOSE-02	SOLID	<u>1.5</u>	<u>1.5</u>	<u>N/A</u>	mrads/hour
98000436	BY-105 (2)	9 SAMPLE	S98T002605 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>5.632</u>		g/L
98000436	BY-105 (2)	10 SAMPLE	S98T002605 0		DOSE-02	SOLID	<u>N/A</u>	<u>1.5</u>		mrads/hour
98000436	BY-105 (2)	11 DUP	S98T002605 0 A		ACIDIG01	SOLID	<u>5.632</u>	<u>5.458</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	12 DUP	S98T002605 0		DOSE-02	SOLID	<u>1.5</u>	<u>1.5</u>	<u>N/A</u>	mrads/hour
98000436	BY-105 (2)	13 SAMPLE	S98T002606 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>5.334</u>		g/L
98000436	BY-105 (2)	14 SAMPLE	S98T002606 0		DOSE-02	SOLID	<u>N/A</u>	<u>6</u>		mrads/hour
98000436	BY-105 (2)	15 DUP	S98T002606 0 A		ACIDIG01	SOLID	<u>5.334</u>	<u>5.416</u>	<u>N/A</u>	g/L
98000436	BY-105 (2)	16 DUP	S98T002606 0		DOSE-02	SOLID	<u>6</u>	<u>6</u>	<u>N/A</u>	mrads/hour

598700 2589 → 598700 2604
2590 → 2605
2591 → 2606

Data Entry Comments:

DATA ENTRY 12-2-98

J. H. Wainall

Validated by

Alanastro 12-10-98

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

worklistrpt Version 2.1 05/15/95
10/26/98 10:44

Page: 2

LABCORE Data Entry Template for Worklist# 26823

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26823
Analyst Signature12-1-98
Date_____
Analyst Signature_____
DateData Entry Comments:
_____
_____
_____

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26825

Analyst: LAZ Instrument: ACD01 Book # WPC-2 2.5mls each into
50mls final volume
Method: LA-505-163 Rev/Mod D-D
Worklist Comment: BY-105 FOR ACIDIG01 RTS D.F.=20x

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	1	.050	N/A	g/L
		2 STD-PREP			ACIDIG01	SOLID	20	20	N/A	g/L
98000436	BY-105 (2)	3 SAMPLE	S98T002607 0 A		ACIDIG01	SOLID	N/A	5.15		g/L
98000436	BY-105 (2)	4 SAMPLE	S98T002607 0		DOSE-02	SOLID	N/A	2.5		mrad/hour
98000436	BY-105 (2)	5 DUP	S98T002607 0 A		ACIDIG01	SOLID	5.15	5.006	N/A	g/L
98000436	BY-105 (2)	6 DUP	S98T002607 0		DOSE-02	SOLID	2.5	2.5	N/A	mrad/hour
98000436	BY-105 (2)	7 SAMPLE	S98T002608 0 A		ACIDIG01	SOLID	N/A	5.056		g/L
98000436	BY-105 (2)	8 SAMPLE	S98T002608 0		DOSE-02	SOLID	N/A	144		mrad/hour
98000436	BY-105 (2)	9 DUP	S98T002608 0 A		ACIDIG01	SOLID	5.056	5.17	N/A	g/L
98000436	BY-105 (2)	10 DUP	S98T002608 0		DOSE-02	SOLID	144	144	N/A	mrad/hour
98000436	BY-105 (2)	11 SPK	S98T002608 0 A		ACIDIG01	SOLID	1	5.3	N/A	g/L
98000436	BY-105 (2)	12 SPK	S98T002608 0		DOSE-02	SOLID	144	174	N/A	mrad/hour
98000436	BY-105 (2)	13 SAMPLE	S98T002609 0 A		ACIDIG01	SOLID	N/A	4.916		g/L
98000436	BY-105 (2)	14 SAMPLE	S98T002609 0		DOSE-02	SOLID	N/A	233		mrad/hour
98000436	BY-105 (2)	15 DUP	S98T002609 0 A		ACIDIG01	SOLID	4.916	5.23	N/A	g/L
98000436	BY-105 (2)	16 DUP	S98T002609 0		DOSE-02	SOLID	233	204	N/A	mrad/hour

S98T002592 → S98T002607
2593 → 2608
2594 → 2609

Data Entry Comments:

DATA ENTRY 12-31-98

John Wenell

Validated by:

Sam M. Pang

01/07/99

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26825

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26825

12/30/98
_____
Analyst Signature   Date

_____
Analyst Signature   Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

worklistrpt Version 2.1 05/15/95
10/26/98 10:47

LABCORE Data Entry Template for Worklist# 26827

Analyst: 2 Instrument: ACD01

Book # WXC-1A 2.5mls each into
WXC-2 50mls total volume
* D.F. = 20x

Method: LA-505-163 Rev/Mod

B-D

Worklist Comment: BY-105 FOR ACIDIG01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	BLNK-PREP				ACIDIG01	SOLID	<u>1</u>	<u>.050</u>	N/A	g/L
		2	STD-PREP				ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
98000436	BY-105 (2)	3	SAMPLE	S98T002713	0	A	ACIDIG01	SOLID	N/A	<u>5.04</u>		g/L
				<u>2520</u>								
98000436	BY-105 (2)	4	SAMPLE	S98T002713	0		DOSE-02	SOLID	N/A	<u>27</u>		mrad/hour
98000436	BY-105 (2)	5	DUP	S98T002713	0	A	ACIDIG01	SOLID	<u>5.04</u>	<u>5.14</u>	N/A	g/L
				<u>.2587</u>								
98000436	BY-105 (2)	6	DUP	S98T002713	0		DOSE-02	SOLID	<u>27</u>	<u>27</u>	N/A	mrad/hour
98000490	BY-105 (2)	7	SAMPLE	S98T002740	0	A	ACIDIG01	SOLID	N/A	<u>4.898</u>		g/L
				<u>.2449</u>								
98000490	BY-105 (2)	8	SAMPLE	S98T002740	0		DOSE-02	SOLID	N/A	<u>1</u>		mrad/hour
98000490	BY-105 (2)	9	DUP	S98T002740	0	A	ACIDIG01	SOLID	<u>4.898</u>	<u>5.154</u>	N/A	g/L
				<u>.2578</u>								
98000490	BY-105 (2)	10	DUP	S98T002740	0		DOSE-02	SOLID	<u>1</u>	<u>1</u>	N/A	mrad/hour
98000490	BY-105 (2)	11	SPK	S98T002740	0	A	ACIDIG01	SOLID	<u>1</u>	<u>5.13</u>	N/A	g/L
				<u>.2545</u>								
98000490	BY-105 (2)	12	SPK	S98T002740	0		DOSE-02	SOLID	<u>1</u>	<u>1</u>	N/A	mrad/hour
98000437	BY-105 (2)	13	SAMPLE	S98T002846	0	A	ACIDIG01	SOLID	N/A	<u>4.904</u>		g/L
				<u>.2452</u>								
98000437	BY-105 (2)	14	SAMPLE	S98T002846	0		DOSE-02	SOLID	N/A	<u>13</u>		mrad/hour
98000437	BY-105 (2)	15	DUP	S98T002846	0	A	ACIDIG01	SOLID	<u>4.904</u>	<u>4.954</u>	N/A	g/L
				<u>.2477</u>								
98000437	BY-105 (2)	16	DUP	S98T002846	0		DOSE-02	SOLID	<u>13</u>	<u>13</u>	N/A	mrad/hour
98000437	BY-105 (2)	17	SPK	S98T002846	0	A	ACIDIG01	SOLID	<u>1</u>	<u>4.95</u>	N/A	g/L
				<u>.2475</u>								
98000437	BY-105 (2)	18	SPK	S98T002846	0		DOSE-02	SOLID	<u>13</u>	<u>13</u>	N/A	mrad/hour

Data Entry Comments:

598T00 2711 → 598T00 2713
2757 → 2740
2840 → 2846

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26827

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26827

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

DATA ENTRY 12-28-98

Validated 12/28/98

John W. Wall

William J. Rand

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26829

Analyst: AMZ

Instrument: ACD01

Book # WKC-1A 2.5 ml each into
WKC-2 50mls final Volume
D.F. = 20x

Method: LA-505-163 Rev/Mod

BO

Worklist Comment: BY-105 FOR ACIDIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>1</u>	<u>0.050</u>	<u>N/A</u>	g/L
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>20</u>	<u>N/A</u>	g/L
98000437	BY-105 (2)	3 SAMPLE	S98T002847	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>4.816</u>		g/L
		<u>.24083</u> → <u>0.050P</u>								
98000437	BY-105 (2)	4 SAMPLE	S98T002847	0	DOSE-02	SOLID	<u>N/A</u>	<u>67</u>		mrad/hour
98000437	BY-105 (2)	5 DUP	S98T002847	0 A	ACIDIG01	SOLID	<u>4.816</u>	<u>5.3</u>	<u>N/A</u>	g/L
		<u>.26505</u> → <u>0.050P</u>								
98000437	BY-105 (2)	6 DUP	S98T002847	0	DOSE-02	SOLID	<u>67</u>	<u>87</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	7 SAMPLE	S98T002926	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>5.05</u>		g/L
		<u>.25253</u> → <u>0.050P</u>								
98000490	BY-105 (2)	8 SAMPLE	S98T002926	0	DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		mrad/hour
98000490	BY-105 (2)	9 DUP	S98T002926	0 A	ACIDIG01	SOLID	<u>5.05</u>	<u>4.99</u>	<u>N/A</u>	g/L
		<u>.24953</u> → <u>0.050P</u>								
98000490	BY-105 (2)	10 DUP	S98T002926	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	11 SAMPLE	S98T002927	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>5.104</u>		g/L
		<u>.25523</u> → <u>0.050P</u>								
98000490	BY-105 (2)	12 SAMPLE	S98T002927	0	DOSE-02	SOLID	<u>N/A</u>	<u>2.5</u>		mrad/hour
98000490	BY-105 (2)	13 DUP	S98T002927	0 A	ACIDIG01	SOLID	<u>5.104</u>	<u>5.248</u>	<u>N/A</u>	g/L
		<u>.26247</u> → <u>0.050P</u>								
98000490	BY-105 (2)	14 DUP	S98T002927	0	DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	<u>N/A</u>	mrad/hour
98000490	BY-105 (2)	15 SPK	S98T002927	0 A	ACIDIG01	SOLID	<u>1</u>	<u>5.21</u>	<u>N/A</u>	g/L
		<u>.26053</u> → <u>0.050P</u>								
98000490	BY-105 (2)	16 SPK	S98T002927	0	DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	<u>N/A</u>	mrad/hour
S98T002847 → S98T002847										
2910 → 2826										
2911 → 2927										

Data Entry Comments:

DATA ENTRY 12-29-98

Validated 12/30/98 Jm Faye

JH Wainell

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26829

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26829

12/28/98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26831

Analyst: For 2 Instrument: ACD01 Book # WHC-1A 2.5mls each into
Method: LA-505-163 Rev/Mod B-0 50mls final volume
Worklist Comment: BY-105 FOR ACIDIG01 RTS D.F. = 20x

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>1</u>	<u>.050</u>	<u>N/A</u>	<u>g/L</u>
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>20</u>	<u>N/A</u>	<u>g/L</u>
98000490	BY-105 (2)	3 SAMPLE	S98T002928 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>4.984</u>		<u>g/L</u>
		<u>.2452</u> → <u>.0500</u>								
98000490	BY-105 (2)	4 SAMPLE	S98T002928 0		DOSE-02	SOLID	<u>N/A</u>	<u>2.5</u>		<u>mrads/hour</u>
98000490	BY-105 (2)	5 DUP	S98T002928 0 A		ACIDIG01	SOLID	<u>4.984</u>	<u>5.264</u>	<u>N/A</u>	<u>g/L</u>
		<u>.2432</u> → <u>.0500</u>								
98000490	BY-105 (2)	6 DUP	S98T002928 0		DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	<u>N/A</u>	<u>mrads/hour</u>
98000490	BY-105 (2)	7 SPK	S98T002928 0 A		ACIDIG01	SOLID	<u>1</u>	<u>5.172</u>	<u>N/A</u>	<u>g/L</u>
		<u>.2586</u> → <u>.0500</u>								
98000490	BY-105 (2)	8 SPK	S98T002928 0		DOSE-02	SOLID	<u>2.5</u>	<u>2.5</u>	<u>N/A</u>	<u>mrads/hour</u>
98000490	BY-105 (2)	9 SAMPLE	S98T002929 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>5.084</u>		<u>g/L</u>
		<u>.2542</u> → <u>.0500</u>								
98000490	BY-105 (2)	10 SAMPLE	S98T002929 0		DOSE-02	SOLID	<u>N/A</u>	<u>2</u>		<u>mrads/hour</u>
98000490	BY-105 (2)	11 DUP	S98T002929 0 A		ACIDIG01	SOLID	<u>5.084</u>	<u>4.92</u>	<u>N/A</u>	<u>g/L</u>
		<u>.2460</u> → <u>.0500</u>								
98000490	BY-105 (2)	12 DUP	S98T002929 0		DOSE-02	SOLID	<u>2</u>	<u>2</u>	<u>N/A</u>	<u>mrads/hour</u>
98000490	BY-105 (2)	13 SAMPLE	S98T002930 0 A		ACIDIG01	SOLID	<u>N/A</u>	<u>5.144</u>		<u>g/L</u>
		<u>.2572</u> → <u>.0500</u>								
98000490	BY-105 (2)	14 SAMPLE	S98T002930 0		DOSE-02	SOLID	<u>N/A</u>	<u>67</u>		<u>mrads/hour</u>
98000490	BY-105 (2)	15 DUP	S98T002930 0 A		ACIDIG01	SOLID	<u>5.144</u>	<u>5.062</u>	<u>N/A</u>	<u>g/L</u>
		<u>.2531</u> → <u>.0500</u>								
98000490	BY-105 (2)	16 DUP	S98T002930 0		DOSE-02	SOLID	<u>67</u>	<u>43</u>	<u>N/A</u>	<u>mrads/hour</u>
98000490	BY-105 (2)	17 SPK	S98T002930 0 A		ACIDIG01	SOLID	<u>1</u>	<u>4.996</u>	<u>N/A</u>	<u>g/L</u>
		<u>.2498</u> → <u>.0500</u>								
98000490	BY-105 (2)	18 SPK	S98T002930 0		DOSE-02	SOLID	<u>67</u>	<u>43</u>	<u>N/A</u>	<u>mrads/hour</u>

Data Entry Comments:

DATA ENTRY 12-28-98

JR Wauell

598700 2912 → 598700 2928
2913 → 2921
2914 → 2920

Validated 12/30/98 smfaga

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26831

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT

Final page for worklist # 26831

12/28/98
_____
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

_____

_____

_____

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26833

Analyst: LMC Instrument: ACD01 Book # NHC-14 2.5mS each into
NHC-2 50 mS final volume
8 D.F. = 20x
 Method: LA-505-163 Rev/Mod B-D
 Worklist Comment: BY-105 FOR ACIDIG01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>1</u>	<u>.050</u>	N/A	g/L
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
98000509	BY-105 (2)	3 SAMPLE	S98T002990	0 A	ACIDIG01	SOLID	N/A	<u>5.156</u>		g/L
		<u>.25177</u> → <u>.0501</u>								
98000509	BY-105 (2)	4 SAMPLE	S98T002990	0	DOSE-02	SOLID	N/A	<u>3</u>		mrad/hour
98000509	BY-105 (2)	5 DUP	S98T002990	0 A	ACIDIG01	SOLID	<u>5.156</u>	<u>5.026</u>	N/A	g/L
		<u>.25177</u> → <u>.0501</u>								
98000509	BY-105 (2)	6 DUP	S98T002990	0	DOSE-02	SOLID	<u>3</u>	<u>3</u>	N/A	mrad/hour
98000509	BY-105 (2)	7 SPK	S98T002990	0 A	ACIDIG01	SOLID	<u>1 N/A</u>	<u>5.274</u>	N/A	g/L
		<u>.26377</u> → <u>.0501</u>								
98000509	BY-105 (2)	8 SPK	S98T002990	0	DOSE-02	SOLID	<u>1 N/A</u>	<u>3</u>	N/A	mrad/hour
98000509	BY-105 (2)	9 SAMPLE	S98T002991	0 A	ACIDIG01	SOLID	N/A	<u>5.044</u>		g/L
		<u>.25227</u> → <u>.0501</u>								
98000509	BY-105 (2)	10 SAMPLE	S98T002991	0	DOSE-02	SOLID	N/A	<u>3</u>		mrad/hour
98000509	BY-105 (2)	11 DUP	S98T002991	0 A	ACIDIG01	SOLID	<u>5.044</u>	<u>4.924</u>	N/A	g/L
		<u>.24627</u> → <u>.0501</u>								
98000509	BY-105 (2)	12 DUP	S98T002991	0	DOSE-02	SOLID	<u>3</u>	<u>3</u>	N/A	mrad/hour
98000509	BY-105 (2)	13 SPK	S98T002991	0 A	ACIDIG01	SOLID	<u>1 N/A</u>	<u>5.31</u>	N/A	g/L
		<u>.26557</u> → <u>.0501</u>								
98000509	BY-105 (2)	14 SPK	S98T002991	0	DOSE-02	SOLID	<u>1 N/A</u>	<u>3</u>	N/A	mrad/hour
98000509	BY-105 (2)	15 SAMPLE	S98T002992	0 A	ACIDIG01	SOLID	N/A	<u>5.118</u>		g/L
		<u>.25577</u> → <u>.0501</u>								
98000509	BY-105 (2)	16 SAMPLE	S98T002992	0	DOSE-02	SOLID	N/A	<u>3</u>		mrad/hour
98000509	BY-105 (2)	17 DUP	S98T002992	0 A	ACIDIG01	SOLID	<u>5.118</u>	<u>5.044</u>	N/A	g/L
		<u>.25227</u> → <u>.0501</u>								
98000509	BY-105 (2)	18 DUP	S98T002992	0	DOSE-02	SOLID	<u>3</u>	<u>3</u>	N/A	mrad/hour

S98T00 2980 → S98T00 2990
 2981 → 2991
 2982 → 2992
 2983 → 2993

Data Entry Comments:

DATA ENTRY 11-19-98
John W. Howell

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

Validated by:
Sam M. Parviz
 11/20/98

LABCORE Data Entry Template for Worklist# 26833

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
98000509	BY-105 (2)	19	SAMPLE	S98T002993	0	A	ACIDIG01	SOLID	N/A	5.014		g/L
				.25077						3		
98000509	BY-105 (2)	20	SAMPLE	S98T002993	0		DOSE-02	SOLID	N/A			mrad/hour
98000509	BY-105 (2)	21	DUP	S98T002993	0	A	ACIDIG01	SOLID	5.014	4972	N/A	g/L
				.24367						3		
98000509	BY-105 (2)	22	DUP	S98T002993	0		DOSE-02	SOLID	3	3	N/A	mrad/hour
98000509	BY-105 (2)	23	SPK	S98T002993	0	A	ACIDIG01	SOLID	N/A	5.15	N/A	g/L
				.25757						3		
98000509	BY-105 (2)	24	SPK	S98T002993	0		DOSE-02	SOLID	N/A	3	N/A	mrad/hour

Final page for worklist # 26833

Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

HPT- Lisa Valdez

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27526

Analyst: ADD Instrument: ACD01 Book # WHC 1A 2.5mLs EACH
 Method: LA-505-163 Rev/Mod B-O Note* TO 50mL Final Volume
 Worklist Comment: SY101, ACIDIG01, tdm BF=20X

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>20</u>	<u>0.05</u>	<u>u/s</u>	<u>12/28/98</u>
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>0.05</u>	<u>N/A</u>	<u>g/L</u>
98000571	SY-101	3 SAMPLE	S98T003350	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>5.006</u>		<u>g/L</u>
98000571	SY-101	4 SAMPLE	S98T003350	0	DOSE-02	SOLID	<u>N/A</u>	<u>33</u>		<u>mrad/hour</u>
98000571	SY-101	5 DUP	S98T003350	0 A	ACIDIG01	SOLID	<u>5.006</u>	<u>5.04</u>	<u>N/A</u>	<u>g/L</u>
98000571	SY-101	6 DUP	S98T003350	0	DOSE-02	SOLID	<u>33</u>	<u>33</u>	<u>N/A</u>	<u>mrad/hour</u>
98000571	SY-101	7 SPK	S98T003350	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>5.038</u>	<u>N/A</u>	<u>g/L</u>
98000571	SY-101	8 SPK	S98T003350	0	DOSE-02	SOLID	<u>33</u>	<u>33</u>	<u>N/A</u>	<u>mrad/hour</u>
98000571	SY-101	9 SAMPLE	S98T003351	0 A	ACIDIG01	SOLID	<u>N/A</u>	<u>5.266</u>		<u>g/L</u>
98000571	SY-101	10 SAMPLE	S98T003351	0	DOSE-02	SOLID	<u>N/A</u>	<u>33</u>		<u>mrad/hour</u>
98000571	SY-101	11 DUP	S98T003351	0 A	ACIDIG01	SOLID	<u>5.266</u>	<u>5.314</u>	<u>N/A</u>	<u>g/L</u>
98000571	SY-101	12 DUP	S98T003351	0	DOSE-02	SOLID	<u>33</u>	<u>33</u>	<u>N/A</u>	<u>mrad/hour</u>

Final page for worklist # 27526

Anthony Purnett 12-20-98
 Analyst Signature Date PARENT

Analyst Signature Date

S98T003344 → S98T003350
 S98T003345 → S98T003351

Data Entry Comments:

DATA ENTRY 12-21-98
J. Howell

Validated 12/21/98
John J. Bond

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
 R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27652

Analyst: PJZ Instrument: ACD01

Book # WBC-1A 2.5 ml each in 30
WBC-2 50 ml final volume
* DF. = 20x

Method: LA-505-163 Rev/Mod B-0

Worklist Comment: BY-105 ACIDIG01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 BLNK-PREP			ACIDIG01	SOLID	<u>1</u>	<u>.050</u>	N/A	g/L
		2 STD-PREP			ACIDIG01	SOLID	<u>20</u>	<u>20</u>	N/A	g/L
98000490	BY-105 (2)	3 SAMPLE	S98T003389	0 A	ACIDIG01	SOLID	N/A	<u>4.878</u>		g/L
		<u>.2437g</u> → <u>.050g</u>								
98000490	BY-105 (2)	4 SAMPLE	S98T003389	0	DOSE-02	SOLID	N/A	<u>2</u>		mrad/hour
98000490	BY-105 (2)	5 DUP	S98T003389	0 A	ACIDIG01	SOLID	<u>4.878</u>	<u>5.272</u>	N/A	g/L
		<u>.2636g</u> → <u>.050g</u>								
98000490	BY-105 (2)	6 DUP	S98T003389	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	N/A	mrad/hour
98000490	BY-105 (2)	7 SPK	S98T003389	0 A	ACIDIG01	SOLID	<u>1</u>	<u>5.178</u>	N/A	g/L
		<u>.2589g</u> → <u>.050g</u>								
98000490	BY-105 (2)	8 SPK	S98T003389	0	DOSE-02	SOLID	<u>2</u>	<u>2</u>	N/A	mrad/hour
98000436	BY-105 (2)	9 SAMPLE	S98T003394	0 A	ACIDIG01	SOLID	N/A	<u>5.088</u>		g/L
		<u>.2544g</u> → <u>.050g</u>								
98000436	BY-105 (2)	10 SAMPLE	S98T003394	0	DOSE-02	SOLID	N/A	<u>3</u>		mrad/hour
98000436	BY-105 (2)	11 DUP	S98T003394	0 A	ACIDIG01	SOLID	<u>5.088</u>	<u>4.866</u>	N/A	g/L
		<u>.2433g</u> → <u>.050g</u>								
98000436	BY-105 (2)	12 DUP	S98T003394	0	DOSE-02	SOLID	<u>3</u>	<u>4</u>	N/A	mrad/hour
98000436	BY-105 (2)	13 SPK	S98T003394	0 A	ACIDIG01	SOLID	<u>1</u>	<u>4.892</u>	N/A	g/L
		<u>.2446g</u> → <u>.050g</u>								
98000436	BY-105 (2)	14 SPK	S98T003394	0	DOSE-02	SOLID	<u>3</u>	<u>2</u>	N/A	mrad/hour
98000436	BY-105 (2)	15 SAMPLE	S98T003399	0 A	ACIDIG01	SOLID	N/A	<u>4.986</u>		g/L
		<u>.2493g</u> → <u>.050g</u>								
98000436	BY-105 (2)	16 SAMPLE	S98T003399	0	DOSE-02	SOLID	N/A	<u>202</u>		mrad/hour
98000436	BY-105 (2)	17 DUP	S98T003399	0 A	ACIDIG01	SOLID	<u>4.986</u>	<u>5.212</u>	N/A	g/L
		<u>.2606g</u> → <u>.050g</u>								
98000436	BY-105 (2)	18 DUP	S98T003399	0	DOSE-02	SOLID	<u>202</u>	<u>230</u>	N/A	mrad/hour

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27652

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
98000436	BY-105 (2)	19 SPK	S98T003399	0 A	ACIDIG01	SOLID	1	4.904	N/A	g/L
	24525		0508							
98000436	BY-105 (2)	20 SPK	S98T003399	0	DOSE-02	SOLID	202	262	N/A	mrads/hour

Final page for worklist # 27652

For Colock 12/22/98
Analyst Signature Date

Analyst Signature Date

S98T003386 → 3389
3391 → 3394
3396 → 3399

Data Entry Comments:

DATA ENTRY 12-28-98
John Warrall

Validated 12/28/98

William J. Baird

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

BULK DENSITY WORKSHEETS

Bulk Density Worksheet

Date: 10/2/98Requestor: F. STEENTank: BY-105Core: 246Seg: 11AAuger: NA

Sample ID: _____

Start Time: _____
End Time: _____
Homogenization
Time (Min.): _____Jar#: 14952Jar/Vial Size: 60 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: _____

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative:

SAMPLE TO DRY
FOR BULK DENSITY.
ADH
10/2/98

327

Bulk Density Worksheet

Date: 10-2-98

Requestor: Stan

Tank: B11705
Core: 646
Seg: 2
Auger:
Sample ID: 5917002584

Start Time: _____

End Type:

Homogenization

Time (Min):

Jar#: 121735

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: 737

Final Vol: 10.07 mL

Initial Weight: 17.55 g

Final Weight: 22.22 g

Net Weight: 14.67 g

Sample ID: _____

Appearance/Narrative:

TO DRY FOR
BULK DENSITY
10/2/98 RCH

10/2/98

328

Bulk Density Worksheet

Date: 12/2/98Requestor: F. STEENTank: B4-65Core: 246Seg: 4LHAuger: 2A

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____
Jar#: 13604Jar/Vial Size: 125 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: _____

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative:

SAMPLE TO DRY
 TO DO BULK DENSITY
 12/2/98
 RDA

Bulk Density Worksheet

Date: 10/2/98Requestor: F. STEWTank: B4-105Core: 246Seg: 2HAuger: NA

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

 Jar#: 13599
 Jar/Vial Size: 125 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

 Cone#: 13599
 Final Vol: 10.0 mL
 Initial Weight: 7.49 g
 Final Weight: 24.94 g
 Net Weight: 17.45 g
 Sample ID: _____

Appearance/Narrative:

$$17.45 / 10.0 = 1.74 \text{ g/mL}$$

Bulk Density Worksheet

Date: 10/2/98Requestor: F. STEIN

Tank: BY-105
 Core: 246
 Seg: 9LH
 Auger: M
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min): _____

Jar#: 14740
 Jar/Vial Size: 250 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Cone#: 14746
 Final Vol: _____ mL
 Initial Weight: 7.54 g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative:

4/1/99
4/1/99
 SAMPLE TO STICKY
 TO DO BULK DENSITY

Bulk Density Worksheet

Date: 10/2/98Requestor: F. STEFFENTank: BY-105Core: 246Seg: 9444Auger: AT

Sample ID: _____

Start Time: _____

End Time: _____

Homogenization

Time (Min.): _____

Jar#: 14951Jar/Vial Size: 60 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: 14951

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative:

SAMPLE WAS TO STICKY
DO BULK DENSITY!
ADH 10/2/98

Bulk Density Worksheet

Date: 12-16-98

Requestor: Star

Tank: B4105
 Core: 246 (244)
 Seg: Comp (1, 244)
 Auger: _____
 Sample ID: 981003375

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: _____
 Jar/Vial Size: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Cone#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative: _____

Jar#: _____
 Jar/Vial Size: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Cone#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative: _____

Bulk Density Worksheet

Date: 12-16-98

HN-1663 REV. 0

Investigator: Steven

Tank: 84-15

Core: 246

Seg: Comp 949A

Auger:

Sample ID: 2981003377

Start Time:
End Time:
Homogenization Time (Min):

Jar#:
Jar/Vial Size: mL
Initial Weight: g
Final Weight: g
Net Weight: g

Cone#:
Final Vol: mL
Initial Weight: g
Final Weight: g
Net Weight: g

Sample ID:

Appearance/Narrative:

Start Time:
End Time:
Homogenization Time (Min):

Jar#:
Jar/Vial Size: mL
Initial Weight: g
Final Weight: g
Net Weight: g

Cone#:
Final Vol: mL
Initial Weight: g
Final Weight: g
Net Weight: g

Sample ID:

Appearance/Narrative:

12-16-98

Material too dry / did not perform bulk density.

Bulk Density Worksheet

Date: 10/2/98Requestor: Steers

Tank: B4105
 Core: 246R
 Seg: 8R L.H.
 Auger:
 Sample ID: SPR002838

Start Time:
 End Time:
 Homogenization
 Time (Min.):

Jar# 38
 Jar/Vial Size: mL
 Initial Weight: g
 Final Weight: g
 Net Weight: g

Core#: 38
 Final Vol: 11.00 mL
 Initial Weight: 7.50 g
 Final Weight: 27.51 g
 Net Weight: 20.01 g
 Sample ID: SPR002838

Appearance/Narrative:

Red/Brown

20.01
 11.00

1.82

1.82

Bulk Density Worksheet

Date: 10/2/98Requestor: SteenTank: B705Core: 246 RSeg: 9R LHAuger: ---Sample ID: 59870014758
 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): 6

 Jar#: 58
 Jar/Vial Size: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

 Cone#: 58
 Final Vol: 9.50 mL
 Initial Weight: 7.51 g
 Final Weight: 24.35 g
 Net Weight: 16.84 g
 Sample ID: 59870014758

 Appearance/Narrative: Red/Brown 1.77

~~24.35~~
~~16.84~~ = ~~1.77~~
 9.50

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEW

Tank: B4 105
 Core: 250
 Seg: 3LH
 Auger: NA
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar# 13601
 Jar/ Trial Size: 125 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Core#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative:

337

SAMPLE WTS TOO
 DRY. 204 10/13/98

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEEN

Tank: BY 405
 Core: 250
 Seg: 44
 Auger: NOT
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: 13600
 Jar/Vial Size: 125 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Core#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative: _____

SAMPLE WAS TOO
 DRY FOR 10/13/98

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEDTank: BY-105Core: C250Seg: SL1Auger: N

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homeogenization
 Time (Min.): _____
Jar#: 14744Jar/Vial Size: 250 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Core#: _____

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

 SAMPLE TOO DRY.
 PSH 10/13/98

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEWTank: BY-05Core: 250Seg: GLVAuger: A4

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____
Jar #: 14738Jar/Vial Size: 250 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Core #: _____

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

 TSS. Very to perform analysis
 3/8/99

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEENTank: BY 105Core: 250Seg: 7 L^HAuger: NA

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

 Jar# 14965
 Jar/ial Size: 60 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

 Cone#: _____
 Final Vol: 10.5 mL
 Initial Weight: 7.51 g
 Final Weight: 27.34 g
 Net Weight: 19.83 g
 Sample ID: _____

Appearance/Narrative: _____

$$19.83 / 10.5 = 1.89 \text{ g/mL}$$

Bulk Density Worksheet

Date: 10/16/98Requestor: F. STEENTank: BY-15Core: 252Seg: 98LHAuger: N/A

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____
Jar #: 14992Jar/Vial Size: 60 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone #: _____

Final Vol: _____ mL

Initial Weight: 7.52 g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

_____

_____

 SAMPLE TOO DRY
 TO DO BULK DENSITY!
 10/16/98 CWT

Bulk Density Worksheet

HNF-1663 REV. 0

Requestor: _____

Date: _____

Tank: B4105

Core: 750

Seg: Comp (34596)

Auger: _____

Sample ID: 598700380

Tank: _____

Core: _____

Seg: _____

Auger: _____

Sample ID: _____

837899

Start Time: _____
End Time: _____
Homogenization Time (min): _____

Start Time: _____
End Time: _____
Homogenization Time (min): _____

Jar#:

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Jar#:

Jar/Vial Size: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#:

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Cone#:

Final Vol: _____ mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____

Appearance/Narrative: _____

Bulk Density Worksheet

Date: 10/5/98Requestor: F. SteenTank: BY-105Core: C251Seg: 2 LHAuger: NA

Sample ID: _____

 Start Time: _____
 End To: _____
 Homogenization
 Time (Min.): _____
Jar# 14982Jar/Vial Size: 60 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Core#: 82

Final Vol: _____ mL

Initial Weight: 7.58 gFinal Weight: 21.60 g

Net Weight: _____ g

Sample ID: _____

Appearance/Narrative: _____
 _____
 _____

Added 10.12.98

Sample is too dry
 still - it didn't
 pack down very good
 unable to do

Bulk Density Worksheet

Date: 10/11/98Requestor: F. STEEN

Tank: 2405
 Core: 257
 Seg: 44
 Auger: _____
 Sample ID: 598T00 2977

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: 14897
 Jar/Vial Size: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Cone#: _____
 Final Vol: 10.5 mL
 Initial Weight: 7.47 g
 Final Weight: 27.59 g
 Net Weight: 7.47 g
 Sample ID: 598T00 2977

Appearance/Narrative:

$$20.12 / 10.5 = 1.92 \text{ g/mL}$$

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEEN

Tank: BY-5
 Core: 251
 Seg: 444
 Auger: NA
 Sample ID: _____

Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____

Jar#: 14984
 Jar/Vial Size: 60 mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g

Core#: _____
 Final Vol: _____ mL
 Initial Weight: _____ g
 Final Weight: _____ g
 Net Weight: _____ g
 Sample ID: _____

Appearance/Narrative:

SAMPLE WAS TOO
 DRY TO DO BULK DENSITY.
 10/13/98

Bulk Density Worksheet

Date: 10/13/98Requestor: F. STEENTank: B4-25Core: 251Seg: SLVAuger: N4

Sample ID: _____

 Start Time: _____
 End Time: _____
 Homogenization
 Time (Min.): _____
Jar#: 14793Jar/Vial Size: 250 mL

Initial Weight: _____ g

Final Weight: _____ g

Net Weight: _____ g

Cone#: _____

Final Vol: 10.75 mLInitial Weight: 7.48 gFinal Weight: 30.90 gNet Weight: 23.42 g

Sample ID: _____

Appearance/Narrative:

23.42 / 10.75 = 2.18 g/mL

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 3 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 693 THRU 1095)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	HYDROXIDE ANALYSIS, OH WORKLIST, AMMONIA ANALYSIS, NH3 WORKLIST, CHROMIUM VI ANALYSIS, CRVI WORKLIST		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955, 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH₃)

NH ₃ Worklist # 26728 (2806, 2807)	787
NH ₃ Worklist # 26730 (2955)	812
NH ₃ Worklist # 26842 (3002, 3003)	826
NH ₃ Worklist # 26882 (2956, 2957)	842
NH ₃ Worklist # 26883 (2958, 2959)	859
NH ₃ Worklist # 26885 (2812, 2813)	877
NH ₃ Worklist # 26886 (2938)	893
NH ₃ Worklist # 26887 (3004)	907
NH ₃ Worklist # 26890 (3808, 3809)	921
NH ₃ Worklist # 27640 (3383)	937
NH ₃ Worklist # 27835 (3005)	950
NH ₃ Worklist # 27849 (2811, 2810)	962
NH ₃ Worklist # 28068 (2391)	980
NH ₃ Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)

IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26731

Analyst: rwk

Instrument: PH01

Book# _____

Method: _____ Rev/Mod D-1

Worklist Comment: BY-105 FOR OH-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	OH-01	LIQUID	1.72e4	1.75e+4	101.744 % Recovery	
2 SAMPLE	S98T002392 0	OH-01	LIQUID	N/A	< 63	62.500 ug/ml	
3 DUP	S98T002392 0	OH-01	LIQUID	< 63	< 63	RPD	
4 SAMPLE	S98T002822 0	OH-01	LIQUID	N/A	8.19e+2	62.500 ug/mL	
5 DUP	S98T002822 0	OH-01	LIQUID	8.19e+2	8.63e+2	5.232 RPD	
6 SAMPLE	S98T002823 0	OH-01	LIQUID	N/A	9.85e+2	62.500 ug/mL	
7 DUP	S98T002823 0	OH-01	LIQUID	9.85e+2	1.06e+3	7.335 RPD	
8 SAMPLE	S98T002824 0	OH-01	LIQUID	N/A	1.13e+3	62.500 ug/mL	
9 DUP	S98T002824 0	OH-01	LIQUID	1.13e+3	1.06e+3	6.393 RPD	
10 SAMPLE	S98T002825 0	OH-01	LIQUID	N/A	1.03e+3	62.500 ug/mL	
11 DUP	S98T002825 0	OH-01	LIQUID	1.03e+3	1.05e+3	1.923 RPD	

Final page for worklist# 26731

Analyst Signature _____ Date _____

Kim Thomas 11-12-98
Analyst Signature Date

[Signature] 11/17/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26731

Analyst: MM Instrument: PH01 Book # 11178

Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY-105 FOR OH-01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				OH-01	LIQUID			N/A	ug/mL
98000436	BY-105 (2)	2	SAMPLE	S98T002392	0		OH-01	LIQUID	N/A			ug/mL
98000436	BY-105 (2)	3	DUP	S98T002392	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	4	SAMPLE	S98T002822	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	5	DUP	S98T002822	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	6	SAMPLE	S98T002823	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	7	DUP	S98T002823	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	8	SAMPLE	S98T002824	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	9	DUP	S98T002824	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	10	SAMPLE	S98T002825	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	11	DUP	S98T002825	0		OH-01	LIQUID			N/A	ug/mL

Final page for worklist # 26731

Analyst Signature Not A Sign Date 11/4/98

Analyst Signature _____ Date _____

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

BLK.

calibration data
 date 98-11-03
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.cal. 24.0 °C
 slope(rel) 1.005
 U(as) 33.7 mV
 electr. input 1
 =====

date 98-11-03 time 21:14
 GET pH 12 # 229
 Id.#1 0000
 Id.#2 .1934
 pH(init) 5.91
 V/ml pH
 EP1 .006 4.59
 stop volt.reached
 =====

5987002822

Id.#1 2022
 C00= .05ml

date 98-11-03 time 22:44
 GET pH 12 # 232
 Id.#1 2022
 Id.#2 .1934
 pH(init) 11.79
 V/ml pH
 EP1 .498 10.20
 EP2 .731 7.99
 stop volt.reached
 =====

date 98-11-03 time 21:14
 GET pH 12 # 229
 .10ml/div dPH=1/div
 start V .000 ml

date 98-11-03 time 22:25
 GET pH 12 # 230
 Id.#1 0000
 Id.#2 .1934
 pH(init) 11.66
 V/ml pH
 EP1 .266 9.96
 EP2 .332 8.03
 manual stop
 =====

*pH too Low
 to Run*

Id.#1 2392
 C00= 2ml
 5987002392
 date 98-11-03 time 22:35
 GET pH 12 # 231
 Id.#1 2392
 Id.#2 .1934
 pH(init) 8.11
 V/ml pH
 EP1 .816 7.16
 manual stop

date 98-11-03 time 22:35
 GET pH 12 # 231
 .10ml/div dPH=1/div
 start V .000 ml

date 98-11-03 time 22:25
 GET pH 12 # 230
 .10ml/div dPH=1/div
 start V .000 ml

date 98-11-03 time 22:44
 GET pH 12 # 232
 .10ml/div dPH=1/div
 start V .000 ml

598T002822

Id.#1 2822-2
C00= .05ml

date 98-11-03 time 22:51
GET pH 12 # 233
Id.#1 2822-2
Id.#2 .1934
pH(init) 11.85
V/ml pH
EP1 .525 10.06
manual stop
=====

date 98-11-03 time 22:51
GET pH 12 # 233
.10ml/div Δ pH=1/div
start V .000 ml

598T002823

Id.#1 2823
C00= .05ml

date 98-11-03 time 23:01
GET pH 12 # 234
Id.#1 2823
Id.#2 .1934
pH(init) 11.87
V/ml pH
EP1 .599 10.21
manual stop
=====

date 98-11-03 time 23:02
GET pH 12 # 234
.10ml/div Δ pH=1/div
start V .000 ml

598T002823.D4P

Id.#1 2823-2
C00= .05ml

date 98-11-03 time 23:10
GET pH 12 # 235
Id.#1 2823-2
Id.#2 .1934
pH(init) 11.83
V/ml pH
EP1 .642 10.21
stop volt.reached
=====

date 98-11-03 time 23:19
GET pH 12 # 235
.10ml/div Δ pH=1/div
start V .000 ml

598T002824

Id.#1 2824
C00= .05ml

date 98-11-03 time 23:30
GET pH 12 # 236
Id.#1 2824
Id.#2 .1934
pH(init) 11.88
V/ml pH
EP1 .685 10.32
EP2 1.078 8.40
stop volt.reached
=====

date 98-11-03 time 23:35
GET pH 12 # 236
.10ml/div dpH=1/div
start V .000 ml

598T002824 040

Id.#1 2824-2
C00= .05ml

date 98-11-03 time 23:43
GET pH 12 # 237
Id.#1 2824-2
Id.#2 .1934
pH(init) 11.84
V/ml pH
EP1 .642 10.29
EP2 .978 8.80
stop volt.reached
=====

date 98-11-03 time 23:44
GET pH 12 # 237
.10ml/div dpH=1/div
start V .000 ml

598T002825

Id.#1 2825
 C00= .05ml
 date 98-11-03 time 23:56
 GET pH 12 # 238
 Id.#1 2825
 Id.#2 .1934
 pH(init) 11.83

	V/ml	pH
EP1	.628	10.21
EP2	.962	8.34

 manual stop
 =====

date 98-11-03 time 23:56
 GET pH 12 # 238
 .10ml/div Δ pH=1/div
 start V .000 ml

598T002825 04p

Id.#1 2825-2
 C00= .05ml
 date 98-11-04 time 00:02
 GET pH 12 # 239
 Id.#1 2825-2
 Id.#2 .1934
 pH(init) 11.84

	V/ml	pH
EP1	.636	10.20

 manual stop
 =====

date 98-11-04 time 00:04
 GET pH 12 # 239
 .10ml/div Δ pH=1/div
 start V .000 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0)

LAB 3-18-99

		BLK
Type	Sample Size (mL) SS	2.000
BLK	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.006
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	5.80E-04
OH-01	OH in Sample in µg/mL (PPM)	9.86E+00
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
BLK		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		BLK
12:02 AM	Concentration of OH in Sample (Molarity)	5.80E-04
	OH in Sample in µg/mL (PPM)	<63

The Result is < Detection Limit

Data Entry by:

Date: 11/12/98

Approved by:

Date:

Form 211102_1 Rev. 1.3

Page 1 of 1

OH (AUTO) : LA-211-102 (D-0) 2-18-99

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.266
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.03E+00
OH-01	OH in Sample in µg/mL (PPM)	1.75E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
STD		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		STD
12:02 AM	Concentration of OH in Sample (Molarity)	1.03E+00
	OH in Sample in µg/mL (PPM)	1.75E+04

Data Entry by:	Date: 11/12/98
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²¹⁸⁹⁹₂₁₈

		SAM
Type	Sample Size (mL) SS	2.000
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.498
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	4.82E-02
OH-01	OH in Sample in µg/mL (PPM)	8.19E+02
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002822		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		SAM
12:02 AM	Concentration of OH in Sample (Molarity)	4.82E-02
	OH in Sample in µg/mL (PPM)	8.19E+02

Data Entry by:	DA	Date:	11/12/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁵⁻⁹⁹_{MD}

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁵⁻⁹⁹ _{MD}		DUP
Type	Sample Size (mL) SS	2.000
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.525
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	5.08E-02
OH-01	OH in Sample in µg/mL (PPM)	8.63E+02
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002822		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time	DUP	
12:02 AM	Concentration of OH in Sample (Molarity)	5.08E-02
	OH in Sample in µg/mL (PPM)	8.63E+02

Data Entry by:

Date: 11/12/98

Approved by:

Date:

Form 211102_1 Rev. 1.3

Page 1 of 1

702

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) *11/2/98*

		SAM
Type	Sample Size (mL) SS	2.000
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.599
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	5.79E-02
OH-01	OH in Sample in µg/mL (PPM)	9.85E+02

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002823		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		
12:02 AM	Concentration of OH in Sample (Molarity)	5.79E-02
	OH in Sample in µg/mL (PPM)	9.85E+02

Data Entry by:	<i>NA</i>	Date:	11/12/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹ ₂₋₁₈₋₉₈ _{ms}

		DUP
Type	Sample Size (mL) SS	2.000
DUP	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.642
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	6.21E-02
OH-01	OH in Sample in µg/mL (PPM)	1.06E+03
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002823		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		DUP
12:02 AM	Concentration of OH in Sample (Molarity)	6.21E-02
	OH in Sample in µg/mL (PPM)	1.06E+03

Data Entry by:		Date:	11/12/98
Approved by:	<i>NA</i>	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁴⁰ ₂₋₁₃₋₉₉

OH (AUTO) : LA-211-102 (D-0) ¹⁴⁰ ₂₋₁₃₋₉₉		SAM
Type	Sample Size (mL) SS	2.000
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.685
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	6.62E-02
OH-01	OH in Sample in µg/mL (PPM)	1.13E+03
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002824		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		SAM
12:02 AM	Concentration of OH in Sample (Molarity)	6.62E-02
	OH in Sample in µg/mL (PPM)	1.13E+03

Data Entry by:

Date: 11/12/98

Approved by:

Date:

Form 211102_1 Rev. 1.3

Page 1 of 1

705

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) *2-14-99*

		DUP
Type	Sample Size (mL) SS	2.000
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.642
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	6.21E-02
OH-01	OH in Sample in µg/mL (PPM)	1.06E+03
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002824		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		
12:02 AM	Concentration of OH in Sample (Molarity)	6.21E-02
	OH in Sample in µg/mL (PPM)	1.06E+03

Data Entry by:	<i>MA</i>	Date:	11/12/98
Approved by:		Date:	

OH (AUTO) : LA-211-102 (D-0) ¹/₂ + 10⁴⁹ _{2+D}

		SAM
Type	Sample Size (mL) SS	2.000
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.628
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	6.07E-02
OH-01	OH in Sample in µg/mL (PPM)	1.03E+03
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002825		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		
12:02 AM	Concentration of OH in Sample (Molarity)	6.07E-02
	OH in Sample in µg/mL (PPM)	1.03E+03

Data Entry by:

Date: 11/12/98

Approved by:

Date:

Form 211102_1 Rev. 1.3

Page 1 of 1

707

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁸ _{LD}

		DUP
Type	Sample Size (mL) SS	2.000
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.636
26731	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	6.15E-02
OH-01	OH in Sample in µg/mL (PPM)	1.05E+03
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002625		
Instrument Code	Detection Limit (µg/mL)	6.25E+01
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
11/04/98		
Time		DUP
12:02 AM	Concentration of OH in Sample (Molarity)	6.15E-02
	OH in Sample in µg/mL (PPM)	1.05E+03

Data Entry by:	<i>NA</i>	Date:	11/12/98
Approved by:		Date:	

LABCORE Completed Worklist Report for Worklist# 26732

Analyst: slh

Instrument: PH01

Book#: 111N8

Method: LA-211-102

Rev/Mod

D-1

Worklist Comment: BY-105 FOR OH-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	OH-01	LIQUID	1.72e4	1.72e4	100.000 % Recovery	
2 SAMPLE	S98T002826 0	OH-01	LIQUID	N/A	5.14e4	2500.000 ug/mL	
3 DUP	S98T002826 0	OH-01	LIQUID	5.14e4	4.99e4	2.962 RPD	
4 SAMPLE	S98T002827 0	OH-01	LIQUID	N/A	5.46e4	2500.000 ug/mL	
5 DUP	S98T002827 0	OH-01	LIQUID	5.46e4	5.35e4	2.035 RPD	
6 SAMPLE	S98T002828 0	OH-01	LIQUID	N/A	6.15e4	2500.000 ug/mL	
7 DUP	S98T002828 0	OH-01	LIQUID	6.15e4	5.74e4	6.897 RPD	
8 SAMPLE	S98T002829 0	OH-01	LIQUID	N/A	6.10e4	2500.000 ug/mL	
9 DUP	S98T002829 0	OH-01	LIQUID	5.49e4	6.10e4	10.526 RPD	
10 SAMPLE	S98T002940 0	OH-01	LIQUID	N/A	5.08e4	2500.000 ug/mL	
11 DUP	S98T002940 0	OH-01	LIQUID	5.44e4	5.08e4	6.844 RPD	

Final page for worklist# 26732

Analyst Signature

11/11/98
Date

Analyst Signature

11/17/98
Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 26732

Analyst: SLH Instrument: PH01 Book # 111N8

Method: LA-211-102 Rev/Mod D-11 140 3-15-99

Worklist Comment: BY-105 FOR OH-01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	2	SAMPLE	S98T002826	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	3	DUP	S98T002826	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	4	SAMPLE	S98T002827	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	5	DUP	S98T002827	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	6	SAMPLE	S98T002828	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	7	DUP	S98T002828	0		OH-01	LIQUID			N/A	ug/mL
98000437	BY-105 (2)	8	SAMPLE	S98T002829	0		OH-01	LIQUID	N/A			ug/mL
98000437	BY-105 (2)	9	DUP	S98T002829	0		OH-01	LIQUID			N/A	ug/mL
98000490	BY-105 (2)	10	SAMPLE	S98T002940	0		OH-01	LIQUID	N/A			ug/mL
98000490	BY-105 (2)	11	DUP	S98T002940	0		OH-01	LIQUID			N/A	ug/mL

Final page for worklist # 26732

Sandra Hood Boatright
Analyst Signature Date 10-25-98

_____
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

date 98-10-25 time 14:30
 GET pH 12 # 167
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 6.78
 V/ml pH
 EP1 .006 5.35
 stop volt.reached
 =====

date 98-10-25 time 14:30
 GET pH 12 # 167
 .10ml/div Δ pH=1/div
 start V .000 ml

.050 ml
 S98T002826 sam

date 98-10-25 time 15:12
 GET pH 12 # 170
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.29
 V/ml pH
 EP1 .782 10.76
 manual stop
 =====

date 98-10-25 time 15:12
 GET pH 12 # 170
 .10ml/div Δ pH=1/div
 start V .000 ml

.050 ml
 S98T002826 dup

date 98-10-25 time 15:18
 GET pH 12 # 171
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.24
 V/ml pH
 EP1 .759 10.73
 manual stop
 =====

date 98-10-25 time 15:19
 GET pH 12 # 171
 .10ml/div Δ pH=1/div
 start V .000 ml

STP 11128 .050 ml
 date 98-10-25 time 14:35
 GET pH 12 # 168
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 11.73
 V/ml pH
 EP1 .262 10.08
 EP2 .353 7.02
 stop volt.reached
 =====

date 98-10-25 time 14:35
 GET pH 12 # 168
 .10ml/div Δ pH=1/div
 start V .000 ml

.050ml

S98T002827 Sam

date 98-10-25 time 15:28
 GET pH 12 # 172
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.27
 V/ml pH
 EP1 .830 10.80
 manual stop
 =====

date 98-10-25 time 15:28
 GET pH 12 # 172
 .10ml/div Δ pH=1/div
 start V .000 ml

.050ml

S98T002827 dup

date 98-10-25 time 15:37
 GET pH 12 # 173
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.25
 V/ml pH
 EP1 .813 10.75
 manual stop
 =====

date 98-10-25 time 15:38
 GET pH 12 # 173
 .10ml/div Δ pH=1/div
 start V .000 ml

.050ml

S98T002828 Sam

date 98-10-25 time 15:54
 GET pH 12 # 175
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.23
 V/ml pH
 EP1 .936 10.80
 manual stop
 =====

date 98-10-25 time 15:55
 GET pH 12 # 175
 .10ml/div Δ pH=1/div
 start V .000 ml

.050ml

S98T002828 dup

date 98-10-25 time 16:01
 GET pH 12 # 176
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.25
 V/ml pH
 EP1 .873 10.79
 manual stop
 =====

date 98-10-25 time 16:02
 GET pH 12 # 176
 .10ml/div dPH=1/div
 start V .000 ml

.050ml

S98T002829 sam

date 98-10-25 time 16:11
 GET pH 12 # 177
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.26
 V/ml pH
 EP1 .927 10.78
 manual stop
 =====

date 98-10-25 time 16:11
 GET pH 12 # 177
 .10ml/div dPH=1/div
 start V .000 ml

.050ml

S98T002829 dup

date 98-10-25 time 16:41
 GET pH 12 # 179
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.17
 V/ml pH
 EP1 .835 10.74
 manual stop
 =====

date 98-10-25 time 16:41
 GET pH 12 # 179
 .10ml/div dPH=1/div
 start V .000 ml

.050ml
 598T002940 *sm*
 date 98-10-25 time 16:51
 GET pH 12 # 180
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.16
       V/ml pH
 EP1 .772 10.73
 manual stop
 =====

date 98-10-25 time 16:51
 GET pH 12 # 180
 .10ml/div Δ pH=1/div
 start V .000 ml

.050ml
 598T002940 *dup*
 date 98-10-25 time 17:11
 GET pH 12 # 182
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 12.17
       V/ml pH
 EP1 .827 10.73
 manual stop
 =====

date 98-10-25 time 17:11
 GET pH 12 # 182
 .10ml/div Δ pH=1/div
 start V .000 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) 2-18-99

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.262
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.01E+00
OH-01	OH in Sample in µg/mL (PPM)	1.72E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
STD		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		
04:00 PM	Concentration of OH in Sample (Molarity)	1.01E+00
	OH in Sample in µg/mL (PPM)	1.72E+04

Data Entry by:	Date:	11/10/98
Approved by:	Date:	
Form 211102_1 Rev. 1.3		

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0)		BLK
Type	Sample Size (mL) SS	0.050
BLK	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.006
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.32E-02
OH-01	OH in Sample in µg/mL (PPM)	3.95E+02
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
BLK		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		BLK
04:00 PM	Concentration of OH in Sample (Molarity)	2.32E-02
	OH in Sample in µg/mL (PPM)	<2500

The Result is < Detection Limit

Data Entry by:	Date:	11/10/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ^{MD} ₂₋₁₄₋₉₉

OH (AUTO) : LA-211-102 (D-0) ^{MD} ₂₋₁₄₋₉₉		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.782
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.02E+00
OH-01	OH in Sample in µg/mL (PPM)	5.14E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002826		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
04:00 PM	Concentration of OH in Sample (Molarity)	3.02E+00
	OH in Sample in µg/mL (PPM)	5.14E+04

Data Entry by:		Date:	11/10/98
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-02449)

OH (AUTO) : LA-211-102 (D-02449)		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.759
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.94E+00
OH-01	OH in Sample in µg/mL (PPM)	4.99E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002826		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
04:00 PM	Concentration of OH in Sample (Molarity)	2.94E+00
	OH in Sample in µg/mL (PPM)	4.99E+04

Data Entry by:	NA	Date:	11/10/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁻¹¹⁻⁹⁹ _{LAB}

OH (AUTO) : LA-211-102 (D-0) ¹⁻¹¹⁻⁹⁹ _{LAB}		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.830
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.21E+00
OH-01	OH in Sample in µg/mL (PPM)	5.46E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002827		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
04:00 PM	Concentration of OH in Sample (Molarity)	3.21E+00
	OH in Sample in µg/mL (PPM)	5.46E+04

Data Entry by:		Date:	11/10/98
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-~~1~~12-18-98)
LAD

OH (AUTO) : LA-211-102 (D- ~~1~~ 12-18-98) LAD		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.813
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.14E+00
OH-01	OH in Sample in µg/mL (PPM)	5.35E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002827		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
04:00 PM	Concentration of OH in Sample (Molarity)	3.14E+00
	OH in Sample in µg/mL (PPM)	5.35E+04

Data Entry by:	NA	Date:	11/11/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ^{1/21/99} _{MSD}

		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.936
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.62E+00
OH-01	OH in Sample in µg/mL (PPM)	6.15E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002828		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
04:00 PM	Concentration of OH in Sample (Molarity)	3.62E+00
	OH in Sample in µg/mL (PPM)	6.15E+04

Data Entry by:	Date: 11/11/98
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁻¹⁵⁻⁹⁸

OH (AUTO) : LA-211-102 (D-0)		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.873
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.38E+00
OH-01	OH in Sample in µg/mL (PPM)	5.74E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002828		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
04:00 PM	Concentration of OH in Sample (Molarity)	3.38E+00
	OH in Sample in µg/mL (PPM)	5.74E+04

Data Entry by:	Date: 11/11/98
Approved by: <i>UA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ^{LAB 2-18-99}

	SAM
Type	Sample Size (mL) SS
SAM	Concentration of HNO3 (Molarity)
Work List	HNO3 Titrant at OH end-point in mL
26732	Dilution Factor DF
Test Code	Concentration of OH in Sample (Molarity)
OH-01	OH in Sample in µg/mL (PPM)

Matrix	
LIQUID	
Sample #	Detection Limit = 125µg / SS * DF
S98T002829	
Instrument Code	Detection Limit (µg/mL)
PH01	2.50E+03
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor
SLH	
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)/((1000000µg/g)/(1000mL/L))
10/25/98	
Time	
04:00 PM	
	SAM
	Concentration of OH in Sample (Molarity)
	OH in Sample in µg/mL (PPM)

Data Entry by:	Date: 11/11/98
Approved by:	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ^{1-18 99} _{AD}

OH (AUTO) : LA-211-102 (D-0) ^{1-18 99} _{AD}		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.835
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.23E+00
OH-01	OH in Sample in µg/mL (PPM)	5.49E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002829		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
04:00 PM	Concentration of OH in Sample (Molarity)	3.23E+00
	OH in Sample in µg/mL (PPM)	5.49E+04

Data Entry by:	<i>NA</i>	Date:	11/11/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁹⁻⁹⁹ _{LAJ}		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.772
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.99E+00
OH-01	OH in Sample in µg/mL (PPM)	5.08E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002840 ²⁻¹⁹⁻⁹⁹		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
04:00 PM	Concentration of OH in Sample (Molarity)	2.99E+00
	OH in Sample in µg/mL (PPM)	5.08E+04

Data Entry by:	Date:	11/11/98
Approved by:	Date:	
Form 211102_1 Rev. 1.3		

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁰⁻¹⁸⁻⁹⁹ _{END}

		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.827
26732	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.20E+00
OH-01	OH in Sample in µg/mL (PPM)	5.44E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002840 ¹⁰⁻¹⁸⁻⁹⁹		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
SLH		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		
04:00 PM		
		DUP
Concentration of OH in Sample (Molarity)		3.20E+00
OH in Sample in µg/mL (PPM)		5.44E+04

Data Entry by:	Date: 11/11/98
Approved by: <i>NA</i>	Date:

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 26733

Analyst: rwk

Instrument: PH01

Book#: 111N8

Method: LA-211-I02 Rev/Mod _____

Worklist Comment: BY-105 FOR OH-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	OH-01	LIQUID	1.72e4	1.74e+4	101.163 % Recovery	
2 SAMPLE	S98T002965 0	OH-01	LIQUID	N/A	3.79e+4	2500.000 ug/mL	
3 DUP	S98T002965 0	OH-01	LIQUID	3.79e+4	3.90e+4	2.861 RPD	
4 SAMPLE	S98T002966 0	OH-01	LIQUID	N/A	3.88e+4	5000.000 ug/mL	
5 DUP	S98T002966 0	OH-01	LIQUID	3.88e+4	4.09e+4 <i>4.09e+4 ug/mL 11/17/98</i>	106.719 RPD	
6 SAMPLE	S98T002967 0	OH-01	LIQUID	N/A	4.89e+4	5000.000 ug/mL	
7 DUP	S98T002967 0	OH-01	LIQUID	4.89e+4	4.93e+4	0.815 RPD	
8 SAMPLE	S98T002968 0	OH-01	LIQUID	N/A	5.10e+4	5000.000 ug/mL	
9 DUP	S98T002968 0	OH-01	LIQUID	5.10e+4	4.72e+4	7.739 RPD	
10 SAMPLE	S98T002969 0	OH-01	LIQUID	N/A	4.77e+4	5000.000 ug/mL	
11 DUP	S98T002969 0	OH-01	LIQUID	4.77e+4	5.76e+4	18.803 RPD	

Final page for worklist# 26733

Analyst Signature _____ Date _____

Kim Thomas 11/12-98
Analyst Signature _____ Date _____

[Signature] 11/17/98
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 26733

Analyst: RL Instrument: PH01 Book # 11178

Method: LA-211-102 Rev/Mod _____

Worklist Comment: BY-105 FOR OH-01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				OH-01	LIQUID			N/A	ug/mL
98000509	BY-105 (2)	2	SAMPLE	S98T002965	0		OH-01	LIQUID	N/A			ug/mL
98000509	BY-105 (2)	3	DUP	S98T002965	0		OH-01	LIQUID			N/A	ug/mL
98000509	BY-105 (2)	4	SAMPLE	S98T002966	0		OH-01	LIQUID	N/A			ug/mL
98000509	BY-105 (2)	5	DUP	S98T002966	0		OH-01	LIQUID			N/A	ug/mL
98000509	BY-105 (2)	6	SAMPLE	S98T002967	0		OH-01	LIQUID	N/A			ug/mL
98000509	BY-105 (2)	7	DUP	S98T002967	0		OH-01	LIQUID			N/A	ug/mL
98000509	BY-105 (2)	8	SAMPLE	S98T002968	0		OH-01	LIQUID	N/A			ug/mL
98000509	BY-105 (2)	9	DUP	S98T002968	0		OH-01	LIQUID			N/A	ug/mL
98000509	BY-105 (2)	10	SAMPLE	S98T002969	0		OH-01	LIQUID	N/A			ug/mL
98000509	BY-105 (2)	11	DUP	S98T002969	0		OH-01	LIQUID			N/A	ug/mL

Final page for worklist # 26733

MLC/Km 10/25/98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

calibration data
 date 98-10-25
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.cal. 23.5 °C
 slope(rel) 1.014
 U(as) 26.2 mV
 electr. input 1
 =====

BLK

date 98-10-25 time 20:43
 GET pH 12 # 183
 Id.#1 2796-2
 Id.#2 .1934
 pH(init) 6.17
 V/ml pH
 EP1 .006 4.93
 stop volt.reached
 =====

SFA

Id.#1 0000
 C00= .05ml
 date 98-10-25 time 20:49
 GET pH 12 # 184
 Id.#1 0000
 Id.#2 .1934
 pH(init) 11.59
 V/ml pH
 EP1 .264 9.69
 stop volt.reached
 =====

date 98-10-25 time 20:51
 GET pH 12 # 184
 .10ml/div Δ pH=1/div
 start V .000 ml

5987002965
LAM.

Id.#1 2965
 C00= .05ml
 date 98-10-25 time 21:21
 GET pH 12 # 185
 Id.#1 2965
 Id.#2 .1934
 pH(init) 11.95
 V/ml pH
 EP1 .576 10.28
 EP2 .880 8.28
 manual stop
 =====

date 98-10-25 time 21:21
 GET pH 12 # 185
 .10ml/div Δ pH=1/div
 start V .000 ml

5987002965
BLP

Id.#1 2965-2
 C00= .05ml
 date 98-10-25 time 21:30
 GET pH 12 # 186
 Id.#1 2965-2
 Id.#2 .1934
 pH(init) 11.96
 V/ml pH
 EP1 .593 10.32
 manual stop
 =====

date 98-10-25 time 21:30
 GET pH 12 # 186
 .10ml/div Δ pH=1/div
 start V .000 ml

598T002966

Id.#1 2966
 C00= .025ml
 date 98-10-25 time 21:37
 GET pH 12 # 187
 Id.#1 2966
 Id.#2 .1934
 pH(init) 11.68
 V/ml pH
 EP1 .295 10.04
 stop volt.reached
 =====

date 98-10-25 time 21:38
 GET pH 12 # 187
 .10ml/div Δ pH=1/div
 start V .000 ml

598T002966

pH
 Id.#1 2966-2
 C00= .025ml
 date 98-10-25 time 21:42
 GET pH 12 # 188
 Id.#1 2966-2
 Id.#2 .1934
 pH(init) 11.66
 V/ml pH
 EP1 .304 10.10
 manual stop
 =====

date 98-10-25 time 21:42
 GET pH 12 # 188
 .10ml/div Δ pH=1/div
 start V .000 ml

598T002967

Id.#1 2967
 C00= .025ml
 date 98-10-25 time 21:52
 GET pH 12 # 189
 Id.#1 2967
 Id.#2 .1934
 pH(init) 11.75
 V/ml pH
 EP1 .372 10.16
 stop volt.reached
 =====

date 98-10-25 time 21:52
 GET pH 12 # 189
 .10ml/div Δ pH=1/div
 start V .000 ml

5987002967 *04P*

Id.#1 2967-2
C00= .025ml

date 98-10-25 time 21:58
GET pH 12 # 190
Id.#1 2967-2
Id.#2 .1934
pH(init) 11.68
V/ml pH
EP1 .375 10.16
manual stop
=====

5987002968

Id.#1 2968
C00= .025ml

date 98-10-25 time 23:57
GET pH 12 # 191
Id.#1 2968
Id.#2 .1934
pH(init) 11.80
V/ml pH
EP1 .388 10.32
stop volt.reached
=====

5987002968 *04P*

Id.#1 2968-2
C00= .025ml

date 98-10-26 time 00:09
GET pH 12 # 192
Id.#1 2968-2
Id.#2 .1934
pH(init) 11.73
V/ml pH
EP1 .359 10.13
stop volt.reached
=====

date 98-10-26 time 00:10
GET pH 12 # 192
.10ml/div ΔpH=1/div
start V .000 ml

598T002969

Id.#1 2969
C00= .025ml

date 98-10-26 time 00:27
GET pH 12 # 193
Id.#1 2969
Id.#2 .1934
pH(init) 11.70
V/ml pH
EP1 .363 10.19
stop volt.reached
=====

date 98-10-26 time 00:27
GET pH 12 # 193
.10ml/div dPH=1/div
start V .000 ml

598T002969
44P

Id.#1 2969-2
C00= .025ml

date 98-10-26 time 00:33
GET pH 12 # 194
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.77
V/ml pH
EP1 .438 10.27
manual stop
=====

date 98-10-26 time 00:33
GET pH 12 # 194
.10ml/div dPH=1/div
start V .000 ml

LD 2-18-99
HNF-1663 REV 0

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹ _{LD}

		BLK
Type	Sample Size (mL) SS	0.050
BLK	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.006
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.32E-02
OH-01	OH in Sample in µg/mL (PPM)	3.95E+02

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
BLK		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*((17.0g/mole)*((1000000µg/g)/(1000mL/L)))	
10/25/98		
Time		BLK
08:49 PM	Concentration of OH in Sample (Molarity)	2.32E-02
	OH in Sample in µg/mL (PPM)	<2500

The Result is < Detection Limit

Data Entry by:		Date:	11/12/98
Approved by:	NA	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁵⁻⁹⁸ (M)

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.264
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.02E+00
OH-01	OH in Sample in µg/mL (PPM)	1.74E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
STD		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		
08:49 PM	Concentration of OH in Sample (Molarity)	1.02E+00
	OH in Sample in µg/mL (PPM)	1.74E+04

Data Entry by:	NA	Date:	11/12/98
Approved by:		Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁴⁰ 2-18-99

		SAM
Type	Sample Size (mL) SS	0.050
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.576
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.23E+00
OH-01	OH in Sample in µg/mL (PPM)	3.79E+04

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002965		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
08:49 PM	Concentration of OH in Sample (Molarity)	2.23E+00
	OH in Sample in µg/mL (PPM)	3.79E+04

Data Entry by:		Date:	11/12/98
Approved by:	<i>NA</i>	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹²⁻¹⁸⁻⁹⁹ _{LAD}

OH (AUTO) : LA-211-102 (D-0) ¹²⁻¹⁸⁻⁹⁹ _{LAD}		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.593
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.29E+00
OH-01	OH in Sample in µg/mL (PPM)	3.90E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002965		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
08:49 PM	Concentration of OH in Sample (Molarity)	2.29E+00
	OH in Sample in µg/mL (PPM)	3.90E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	
Form 211102_1 Rev. 1.3		

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-~~0~~)¹²⁻¹⁸⁻⁹⁸ _{LAB}

		SAM
Type	Sample Size (mL) SS	0.025
SAM	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.295
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.28E+00
OH-01	OH in Sample in µg/mL (PPM)	3.88E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002966		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
08:49 PM	Concentration of OH in Sample (Molarity)	2.28E+00
	OH in Sample in µg/mL (PPM)	3.88E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹₂₅₀

		DUP
Type	Sample Size (mL) SS	0.025
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.304
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.35E+00
OH-01	OH in Sample in µg/mL (PPM)	4.00E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002966		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		
08:49 AM	Concentration of OH in Sample (Molarity)	2.35E+00
	OH in Sample in µg/mL (PPM)	4.00E+04

Data Entry by:	Date:	11/17/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹ *WJ*

		SAM
Type	Sample Size (mL) SS	0.025
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.372
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.88E+00
OH-01	OH in Sample in µg/mL (PPM)	4.89E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002967		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
08:49 PM	Concentration of OH in Sample (Molarity)	2.88E+00
	OH in Sample in µg/mL (PPM)	4.89E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹ _{LED}

		DUP
Type	Sample Size (mL) SS	0.025
DUP	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.375
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.90E+00
OH-01	OH in Sample in µg/mL (PPM)	4.93E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002967		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
08:49 PM	Concentration of OH in Sample (Molarity)	2.90E+00
	OH in Sample in µg/mL (PPM)	4.93E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁰⁻¹⁸⁻⁹⁸		SAM
Type	Sample Size (mL) SS	0.025
SAM	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.388
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.00E+00
OH-01	OH in Sample in µg/mL (PPM)	5.10E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002968		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
08:49 PM	Concentration of OH in Sample (Molarity)	3.00E+00
	OH in Sample in µg/mL (PPM)	5.10E+04

Data Entry by:	Date: 11/12/98
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹²⁻¹⁸⁻⁹⁹ ₁₂₋₁₈₋₉₉		DUP
Type	Sample Size (mL) SS	0.025
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.359
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.78E+00
OH-01	OH in Sample in µg/mL (PPM)	4.72E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002968		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		DUP
08:49 PM	Concentration of OH in Sample (Molarity)	2.78E+00
	OH in Sample in µg/mL (PPM)	4.72E+04

Data Entry by:	Date: 11/12/98
Approved by:	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁸⁻⁹⁹ _{LAB}

		SAM
Type	Sample Size (mL) SS	0.025
SAM	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.363
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.81E+00
OH-01	OH in Sample in µg/mL (PPM)	4.77E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002969		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		SAM
08:49 PM	Concentration of OH in Sample (Molarity)	2.81E+00
	OH in Sample in µg/mL (PPM)	4.77E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹⁸⁹⁹ ₂₅

		DUP
Type	Sample Size (mL) SS	0.025
DUP	Concentration of HNO ₃ (Molarity)	0.1934
Work List	HNO ₃ Titrant at OH end-point in mL	0.438
26733	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.39E+00
OH-01	OH in Sample in µg/mL (PPM)	5.76E+04

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T002969		
Instrument Code	Detection Limit (µg/mL)	5.00E+03
PH01		
Analyst	OH Molarity = ((mL HNO ₃)*(M HNO ₃))/Sample Size in mL*Dilution Factor	
RWK		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/25/98		
Time		
08:49 PM	Concentration of OH in Sample (Molarity)	3.39E+00
	OH in Sample in µg/mL (PPM)	5.76E+04

Data Entry by:	Date:	11/12/98
Approved by:	Date:	

LABCORE Completed Worklist Report for Worklist# 26843

Analyst: jds

Instrument: PH01

Book#: 111N8

Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY-105 FOR 0H-01 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		OH-01	LIQUID	1.72e4	1.70e4	98.837 % Recovery	
2 BLNK	0		OH-01	LIQUID	1	<2.5e+3		ug/mL
3 SAMPLE	S98T003010	0	OH-01	LIQUID	N/A	4.16e4	2500.000	ug/mL
4 DUP	S98T003010	0	OH-01	LIQUID	4.16e+4	3.91e+4	6.196	RPD
5 SAMPLE	S98T003011	0	OH-01	LIQUID	N/A	4.33e+4	2500.000	ug/mL
6 DUP	S98T003011	0	OH-01	LIQUID	4.33e+4	4.47e+4	3.182	RPD
7 SAMPLE	S98T003012	0	OH-01	LIQUID	N/A	4.77e+4	2500.000	ug/mL
8 DUP	S98T003012	0	OH-01	LIQUID	4.77e+4	5.02e+4	5.107	RPD
9 SAMPLE	S98T003013	0	OH-01	LIQUID	N/A	4.73e+4	2500.000	ug/mL
10 DUP	S98T003013	0	OH-01	LIQUID	4.73e+4	4.63e+4	2.137	RPD

Final page for worklist# 26843

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 26843

Analyst: JS Instrument: PH01 Book # 111N8A

Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY-105 FOR OH-01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				OH-01	LIQUID	<u>1.72E+4</u>	<u>1.70E+4</u>		
			• B2K							<u>2.84E+2</u>	<u>11-17-98</u>	
98000510	BY-105 (2)	2	SAMPLE	S98T003010	0		OH-01	LIQUID	<u>N/A</u>	<u>4.2500</u>	<u>N/A</u>	ug/mL
										<u>4.16E+4</u>	<u>2.50E+3</u>	
98000510	BY-105 (2)	3	DUP	S98T003010	0		OH-01	LIQUID	<u>4.16E+4</u>	<u>3.91E+4</u>	<u>N/A</u>	ug/mL
98000510	BY-105 (2)	4	SAMPLE	S98T003011	0		OH-01	LIQUID	<u>N/A</u>	<u>4.33E+4</u>	<u>2.5E+3</u>	ug/mL
98000510	BY-105 (2)	5	DUP	S98T003011	0		OH-01	LIQUID	<u>4.33E+4</u>	<u>4.47E+4</u>	<u>N/A</u>	ug/mL
98000510	BY-105 (2)	6	SAMPLE	S98T003012	0		OH-01	LIQUID	<u>N/A</u>	<u>4.77E+4</u>	<u>2.5E+3</u>	ug/mL
98000510	BY-105 (2)	7	DUP	S98T003012	0		OH-01	LIQUID	<u>4.77E+4</u>	<u>5.02E+4</u>	<u>N/A</u>	ug/mL
98000510	BY-105 (2)	8	SAMPLE	S98T003013	0		OH-01	LIQUID	<u>N/A</u>	<u>4.73E+4</u>	<u>2.5E+3</u>	ug/mL
98000510	BY-105 (2)	9	DUP	S98T003013	0		OH-01	LIQUID	<u>4.73E+4</u>	<u>4.63E+4</u>	<u>N/A</u>	ug/mL

Final page for worklist # 26843

JS
Analyst Signature
10/27/98
Date

C. P. Quinn
Analyst Signature
11/16/98
Date

Data Entry Comments:

Batch # 5447

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

STD
H118A

BLANK

3010 STM
.050

date 98-10-27 time 10:46
GET pH 12 # 196
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.55
V/ml pH
EP1 .259 10.18
stop volt.reached
=====

date 98-10-27 time 10:46
GET pH 12 # 196
.10ml/div Δ pH=1/div
start V .000 ml

date 98-10-27 time 10:36
GET pH 12 # 195
Id.#1 2969-2
Id.#2 .1934
pH(init) 6.29
V/ml pH
EP1 .008 5.07
manual stop
=====

date 98-10-27 time 10:36
GET pH 12 # 195
.10ml/div Δ pH=1/div
start V .000 ml

date 98-10-27 time 11:22
GET pH 12 # 197
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.94
V/ml pH
EP1 .632 10.38
manual stop
=====

date 98-10-27 time 11:22
GET pH 12 # 197
.10ml/div Δ pH=1/div
start V .000 ml

3010 DUP
.050

date 98-10-27 time 11:29
GET pH 12 # 198
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.92
V/ml pH
EP1 .594 10.34
manual stop
=====

date 98-10-27 time 11:31
GET pH 12 # 198
.10ml/div dPH=1/div
start V .000 ml

30115AM .050

date 98-10-27 time 11:39
GET pH 12 # 199
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.89
V/ml pH
EP1 .659 10.42
manual stop
=====

date 98-10-27 time 11:39
GET pH 12 # 199
.10ml/div dPH=1/div
start V .000 ml

3611 DUP .050

date 98-10-27 time 11:46
GET pH 12 # 200
Id.#1 2969-2
Id.#2 .1934
pH(init) 11.96
V/ml pH
EP1 .688 10.43
manual stop
=====

date 98-10-27 time 11:46
GET pH 12 # 200
.10ml/div dPH=1/div
start V .000 ml

3012 SKM .550
 date 98-10-27 time 11:55
 GET pH 12 # 201
 Id.#1 2969-2
 Id.#2 .1934
 pH(init) 11.95
 V/ml pH
 EP1 .725 10.46
 manual stop
 =====

302
 date 98-10-27 time 11:57
 GET pH 12 # 201
 .10ml/div Δ pH=1/div
 start V .000 ml

date 98-10-27 time 12:03
 GET pH 12 # 202
 Id.#1 2969-2
 Id.#2 .1934
 pH(init) 11.98
 V/ml pH
 EP1 .764 10.48
 manual stop
 =====

3012 DUP .050

date 98-10-27 time 12:03
 GET pH 12 # 202
 .10ml/div Δ pH=1/div
 start V .000 ml

3013 SKM .050
 date 98-10-27 time 12:10
 GET pH 12 # 203
 Id.#1 2969-2
 Id.#2 .1934
 pH(init) 11.91
 V/ml pH
 EP1 .719 10.41
 manual stop
 =====

date 98-10-27 time 12:10
 GET pH 12 # 203
 .10ml/div Δ pH=1/div
 start V .000 ml

date 98-10-27 time 12:17
GET pH 12 # 204
Id.#1 .2969-2
Id.#2 .1934
pH(init) 11.86
V/ml pH
EP1 .704 10.41
manual stop
=====

30/3D₀R .050

date 98-10-27 time 12:18
GET pH 12 # 204
.10ml/div dpH=1/div
start V .000 ml

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) <i>2-18-99 LND</i>		BLANK
Type	Sample Size (mL) SS	0.050
BLANK	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.008
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	3.09E-02
OH-01	OH in Sample in µg/mL (PPM)	5.26E+02
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
BLANK		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		
11:50 AM	Concentration of OH in Sample (Molarity)	3.09E-02
	OH in Sample in µg/mL (PPM)	<2500

The Result is < Detection Limit

Data Entry by:	Date: 11/16/98
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D- ~~0~~) ^{LAD} 2-15-99		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.259
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	1.00E+00
OH-01	OH in Sample in µg/mL (PPM)	1.70E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
STD		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
115 JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		STD
11:50 AM	Concentration of OH in Sample (Molarity)	1.00E+00
	OH in Sample in µg/mL (PPM)	1.70E+04

Data Entry by:	Date:	11/16/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹ _{LD}		SAMPLE
Type	Sample Size (mL) SS	0.050
SAMPLE	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.632
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.44E+00
OH-01	OH in Sample in µg/mL (PPM)	4.16E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003010		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
JIS JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		SAMPLE
11:50 AM	Concentration of OH in Sample (Molarity)	2.44E+00
	OH in Sample in µg/mL (PPM)	4.16E+04

Data Entry by:	Date:	11/16/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) 2-18-99 LAD		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.594
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.30E+00
OH-01	OH in Sample in µg/mL (PPM)	3.91E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003010		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
JIS JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		DUP
11:50 AM	Concentration of OH in Sample (Molarity)	2.30E+00
	OH in Sample in µg/mL (PPM)	3.91E+04

Data Entry by:	Date: 11/16/98
Approved by: <i>AA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) 2-1899
LA-1

		SAMPLE
Type	Sample Size (mL) SS	0.050
SAMPLE	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.659
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.55E+00
OH-01	OH in Sample in µg/mL (PPM)	4.33E+04

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003011		
Instrument Code	Detection Limit (µg/mL)	2.50E+03

PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
JIS JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		

Time		
11:50 AM	Concentration of OH in Sample (Molarity)	2.55E+00
	OH in Sample in µg/mL (PPM)	4.33E+04

Data Entry by:	Date:	11/16/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-~~0~~) ²⁻¹⁵⁻⁹⁹ _{MSD}

		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.680
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.63E+00
OH-01	OH in Sample in µg/mL (PPM)	4.47E+04

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003011		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
MS JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		
11:50 AM	Concentration of OH in Sample (Molarity)	2.63E+00
	OH in Sample in µg/mL (PPM)	4.47E+04

Data Entry by:	Date:	11/16/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ²⁻¹⁸⁻⁹⁹ _{LAD}		SAMPLE
Type	Sample Size (mL) SS	0.050
SAMPLE	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.725
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.80E+00
OH-01	OH in Sample in µg/mL (PPM)	4.77E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003012		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
JIS JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		SAMPLE
11:50 AM	Concentration of OH in Sample (Molarity)	2.80E+00
	OH in Sample in µg/mL (PPM)	4.77E+04

Data Entry by:	Date: 11/16/98
Approved by: <i>NA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) ¹³⁹⁹ _{LD}

		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.764
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.96E+00
OH-01	OH in Sample in µg/mL (PPM)	5.02E+04

Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003012		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
118 JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		DUP
11:50 AM	Concentration of OH in Sample (Molarity)	2.96E+00
	OH in Sample in µg/mL (PPM)	5.02E+04

Data Entry by:	Date: 11/16/98
Approved by: <i>MA</i>	Date:

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) 2-18-99 LAD		SAMPLE
Type	Sample Size (mL) SS	0.050
SAMPLE	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.719
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.78E+00
OH-01	OH in Sample in µg/mL (PPM)	4.73E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003013		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
J18 JDS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time		SAMPLE
11:50 AM	Concentration of OH in Sample (Molarity)	2.78E+00
	OH in Sample in µg/mL (PPM)	4.73E+04

Data Entry by:	Date:	11/16/98
Approved by:	Date:	

PLACE ANALYTICAL CARD IN BOX BELOW OR ATTACH TRAVELER

OH (AUTO) : LA-211-102 (D-0) <i>2-18-99 LAD</i>		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of HNO3 (Molarity)	0.1934
Work List	HNO3 Titrant at OH end-point in mL	0.704
26843	Dilution Factor DF	1
Test Code	Concentration of OH in Sample (Molarity)	2.72E+00
OH-01	OH in Sample in µg/mL (PPM)	4.63E+04
Matrix		
LIQUID		
Sample #	Detection Limit = 125µg / SS * DF	
S98T003013		
Instrument Code	Detection Limit (µg/mL)	2.50E+03
PH01		
Analyst	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL*Dilution Factor	
<i>90</i> JS JS		
Date	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*((1000000µg/g)/(1000mL/L))	
10/27/98		
Time	DUP	
11:50 AM	Concentration of OH in Sample (Molarity)	2.72E+00
	OH in Sample in µg/mL (PPM)	4.63E+04

Data Entry by:	Date: 11/16/98
Approved by: <i>NA</i>	Date:

LABCORE Completed Worklist Report for Worklist# 27642

Analyst: pjm

Instrument: PH01

Book#: 111N8

Method: LA-211-102 Rev/Mod _____

Worklist Comment: BY-105 OH-01 cjo

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD	0		OH-01	LIQUID	1.72e4	1.75e+4	101.744 % Recovery	
2	BLNK	0		OH-01	LIQUID	1	<2500		ug/mL
3	SAMPLE	998T003385	0	OH-01	LIQUID	N/A	4.64e+4	125.000	ug/mL
4	DUP	998T003385	0	OH-01	LIQUID	4.64e+4	4.55e+4	1.959 RPD	

Final page for worklist# 27642

Analyst Signature _____ Date _____

C. J. Pinew 12/28/98
Analyst Signature Date

[Signature] 12/31/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27642 Analyst: BLW Instrument: PH01 Book # 11128Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY-105 OH-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			OH-01	LIQUID	<u>1.75E+4</u>	<u>N/A</u>	<u>2-27-98</u>	ug/mL
98000437	BY-105 (2)	2 SAMPLE	S98T003385	0	OH-01	LIQUID	<u>N/A</u>	<u>2500</u>	<u>2500</u>	ug/mL
98000437	BY-105 (2)	3 DUP	S98T003385	0	OH-01	LIQUID			<u>N/A</u>	ug/mL

Final page for worklist # 27642.BL McCown 12/20/98
Analyst Signature DateC.D. Owen 12/28/98
Analyst Signature Date

Data Entry Comments:

Blank # 3385Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

calibration data
 date 98-12-20
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.col. 23.5 °C
 slope(rel) .997
 U(as) 35.0 mV
 electr. input 1
 =====

BLANK

date 98-12-20 time 04:38
 GET pH 12 # 322
 Id.#1 0030
 Id.#2 .1934
 pH(init) 6.62
 V/ml pH
 EP1 .008 5.25
 stop volt.reached
 =====

date 98-12-20 time 04:38
 GET pH 12 # 322
 .10ml/div Δ pH=1/div
 start V .000 ml

As McCown 12/20/98
 STD 11118 1050ML

date 98-12-20 time 04:44
 GET pH 12 # 323
 Id.#1 0030
 Id.#2 .1934
 pH(init) 11.58
 V/ml pH
 EP1 .266 9.74
 EP2 .337 7.02
 stop volt.reached
 =====

date 98-12-20 time 04:45
 GET pH 12 # 323
 .10ml/div Δ pH=1/div
 start V .000 ml

5987003385 .050ML
 date 98-12-20 time 05:05
 GET pH 12 # 324
 Id.#1 0030
 Id.#2 .1934
 pH(init) 12.05
 V/ml pH
 EP1 .706 10.50
 EP2 1.138 8.22
 EP3 1.187 7.61
 EP4 1.244 6.90
 stop volt.reached
 =====

date 98-12-20 time 05:06
 GET pH 12 # 324
 .10ml/div Δ pH=1/div
 start V .000 ml

3385 DUP .050ML
 date 98-12-20 time 05:17
 GET pH 12 # 325
 Id.#1 0030
 Id.#2 .1934
 pH(init) 12.04
 V/ml pH
 EP1 .692 10.51
 EP2 1.105 8.46
 EP3 1.176 7.60
 EP4 1.255 6.64
 stop volt.reached
 =====

date 98-12-20 time 05:17
 GET pH 12 # 325
 .10ml/div Δ pH=1/div
 start V .000 ml

OH (AUTO) : LA-211-102 (D-1)

		STANDARD
Type	Sample Size (mL) SS	0.050
STANDARD	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.266
27642	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	1.03E+00
Matrix	Concentration of Sample in PPM	1.75E+04
LIQUID		
Sampler#		
STD	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01		2.50E+03
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
12/20/98	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
04:30 AM		
		STANDARD
Concentration of Sample (MOLARITY)		1.03E+00
Concentration of Sample in PPM		1.75E+04

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		BLANK
Type	Sample Size (mL) SS	0.050
BLANK	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.008
27642	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	3.09E-02
Matrix	Concentration of Sample in PPM	5.26E+02
LIQUID		
Sample #		
BLANK	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01	Detection Limit (PPM)	2.50E+03
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
12/20/98	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
04:30 AM		
		BLANK
	Concentration of Sample (MOLARITY)	3.09E-02
	Concentration of Sample in PPM	<2500
	The Result is < Detection Limit	

OH (AUTO) : LA-211-102 (D-1)

		SAMPLE
Type	Sample Size (mL) SS	0.050
SAMPLE	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.706
27642	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	2.73E+00
Matrix	Concentration of Sample in PPM	4.64E+04
LIQUID		
Sample #		
S98T003385	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01	Detection Limit (PPM)	2.50E+03
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
12/20/98	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
04:30 AM		
		SAMPLE
Concentration of Sample (MOLARITY)		2.73E+00
Concentration of Sample in PPM		4.64E+04

OH (AUTO) : LA-211-102 (D-1)

		DUP
Type	Sample Size (mL) SS	0.050
DUP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.692
27642	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	2.68E+00
Matrix	Concentration of Sample in PPM	4.55E+04
LIQUID		
Sample #		
S98T003385	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01	Detection Limit (PPM)	2.50E+03
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
12/20/98	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
04:30 AM		
		DUP
Concentration of Sample (MOLARITY)		2.68E+00
Concentration of Sample in PPM		4.55E+04

LABCORE Data Entry Template for Worklist# 28070

Analyst: RSW Instrument: PH01 Book # 111 N8

Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY105 FB FOR OH-01 RLC

GROUP	PROJECT	S TYPE	SAMPLE#	R A	TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 STD			OH-01	LIQUID	<u>1.72e4</u>	<u>1.72e4</u>	<u>N/A</u>	ug/mL
98000436	BY-105 (2)	2 SAMPLE	S98T002391	0	OH-01	LIQUID	<u>N/A</u>	<u><125</u>		ug/mL
98000436	BY-105 (2)	3 DUP	S98T002391	0	OH-01	LIQUID	<u><125</u>	<u><125</u>	<u>N/A</u>	ug/mL

Final page for worklist # 28070

RSW 1/27/99
Analyst Signature Date

Mary Jane 2-1-99
Analyst Signature Date

AM/ 2/1/99
REVIEWER SIGNATURE DATE

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

calibration data
 date 99-01-27
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.cal. 23.5 °C
 slope(rel) .996
 U(as) -31.6 mV
 electr. input 1
 =====

date 99-01-27 time 22:30
 GET pH 12 # 1
 Id.#1 255
 Id.#2 .1934
 pH(init) 5.82
 V/ml pH
 EP1 .006 4.90
 manual stop
 =====

Blank

date 99-01-27 time 22:30
 GET pH 12 # 1
 .10ml/div Δ pH=1/div
 start V .000 ml

date 99-01-27 time 22:36
 GET pH 12 # 2
 Id.#1 255
 Id.#2 .1934
 pH(init) 12.02
 V/ml pH
 EP1 .262 10.36
 EP2 .342 7.24
 stop volt.reached
 =====

STD # 111N8 .050M

BSMcLean 1/27/99

date 99-01-27 time 22:38
 GET pH 12 # 2
 .10ml/div Δ pH=1/div
 start V .000 ml

date 99-01-27 time 22:43
 GET pH 12 # 3
 Id.#1 255
 Id.#2 .1934
 pH(init) 6.69
 V/ml pH
 EP1 .007 5.70
 stop volt.reached
 =====

SP8 T002391 2.0ML

date 99-01-27 time 22:43
 GET pH 12 # 3
 .10ml/div Δ pH=1/div
 start V .000 ml

date 99-01-27 time 22:46
 GET pH 12 # 4
 Id.#1 255
 Id.#2 .1934
 pH(init) 7.31
 V/ml pH
 EP1 .008 6.21
 stop volt.reached
 =====

2391 Oap 2.0ML

date 99-01-27 time 22:46
 GET pH 12 # 4
 .10ml/div Δ pH=1/div
 start V .000 ml

OH (AUTO) : LA-211-102 (D-1)

		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.262
28070	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	1.01E+00
Matrix	Concentration of Sample in PPM	1.72E+04
LIQUID		
Sample #		
STD		
Instrument Code		
PH01	Detection Limit (PPM)	
Analyst		
PJM	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
01/27/99	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
10:45 PM		
		STD
	Concentration of Sample (MOLARITY)	1.01E+00
	Concentration of Sample in PPM	1.72E+04

OH (AUTO) : LA-211-102 (D-1)

		BLNK
Type	Sample Size (mL) SS	2.000
BLNK	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.006
28070	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	5.80E-04
Matrix	Concentration of Sample in PPM	9.86E+00
LIQUID		
Sample #		
BLNK	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01	Detection Limit (PPM)	6.25E+01
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
01/27/99	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
10:45 PM		
		BLNK
Concentration of Sample (MOLARITY)		5.80E-04
Concentration of Sample in PPM		<63
The Result is < Detection Limit		

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		SAMPLE
Type	Sample Size (mL) SS	2.000
SAMPLE	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.007
28070	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	6.77E-04
Matrix	Concentration of Sample in PPM	1.15E+01
LIQUID		
Sample #		
S98T002391	Detection Limit =(125µg/SS)*DF	
Instrument Code		
PH01	Detection Limit (PPM)	6.25E+01
Analyst		
PJM	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
01/27/99	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
10:45 PM		
		SAMPLE
	Concentration of Sample (MOLARITY)	6.77E-04
	Concentration of Sample in PPM	<63
	The Result is < Detection Limit	

OH (AUTO) : LA-211-102 (D-1)

		DUP
Type	Sample Size (mL) SS	2.000
DUP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.008
28070	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample (MOLARITY)	7.74E-04
Matrix	Concentration of Sample in PPM	1.32E+01
LIQUID		
Sample #		
S98T002391	Detection Limit $= (125\mu\text{g}/\text{SS}) * \text{DF}$	
Instrument Code		
PH01	Detection Limit (PPM)	6.25E+01
Analyst		
PJM	OH Molarity $= ((\text{mL HNO}_3) * (\text{M HNO}_3)) / \text{Sample Size in mL} * \text{Dilution Factor}$	
Date		
01/27/99	OH in $\mu\text{g}/\text{mL} = (\text{OH MOLARITY}) * (17.0\text{g}/\text{mole}) * (1000000\mu\text{g}/\text{g}) / (1000\text{mL}/\text{L})$	
Time		
10:45 PM		
		DUP
Concentration of Sample (MOLARITY)		7.74E-04
Concentration of Sample in PPM		<63
The Result is < Detection Limit		

LABCORE Completed Worklist Report for Worklist# 28138

Analyst: jjs

Instrument: PH01

Book#: 111N8

Method: LA-211-102 Rev/Mod _____

Worklist Comment: BY105 FOR OH-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	OH-01	SOLID	1.72e4	1.66e4	96.512 % Recovery	
2 BLNK-PREP	0	OH-01	SOLID	1	<24821		ug/g
3 SAMPLE	S98T003390 0	OH-01	SOLID	N/A <	24821	24800.000	ug/g
4 DUP	S98T003390 0	OH-01	SOLID	<24821	<24352		RPD
5 SAMPLE	S98T003395 0	OH-01	SOLID	N/A <	25293	25300.000	ug/g
6 DUP	S98T003395 0	OH-01	SOLID	<25293	<24234		RPD
7 SAMPLE	S98T003400 0	OH-01	SOLID	N/A	4.28e4	24700.000	ug/g
8 DUP	S98T003400 0	OH-01	SOLID	4.28e4	<24645		RPD

Final page for worklist# 28138

Analyst Signature _____

Date _____

Analyst Signature bermuda Date 2-4-99

Reviewer Signature [Signature]

Date 2/9/99

LABCORE Data Entry Template for Worklist# 28138

Analyst: UIS Instrument: PH01 Book # 111NR

Method: LA-211-102 Rev/Mod D-1

Worklist Comment: BY105 FOR OH-01 RLC

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	STD				OH-01	SOLID			N/A	ug/g
		2	BLNK-PREP				OH-01	SOLID			N/A	ug/g
98000490	BY-105 (2)	3	SAMPLE	S98T003390	0		OH-01	SOLID	N/A			ug/g
98000490	BY-105 (2)	4	DUP	S98T003390	0		OH-01	SOLID			N/A	ug/g
98000436	BY-105 (2)	5	SAMPLE	S98T003395	0		OH-01	SOLID	N/A			ug/g
98000436	BY-105 (2)	6	DUP	S98T003395	0		OH-01	SOLID			N/A	ug/g
98000436	BY-105 (2)	7	SAMPLE	S98T003400	0		OH-01	SOLID	N/A			ug/g
98000436	BY-105 (2)	8	DUP	S98T003400	0		OH-01	SOLID			N/A	ug/g

Final page for worklist # 28138

Analyst Signature

02-03-99
Date

_____
Analyst Signature

_____
Date

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

calibration data
 date 99-02-03
 pH(S) 1 7.00
 pH(S) 2 10.00
 t.cal. 24.0 °C
 slope(rel) .992
 U(as) -44.4 mV
 electr. input 1
 =====

81K
 date 99-02-03 time 13:04
 GET pH 12 # 19
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 5.97
 V/ml pH
 EP1 .008 4.74
 stop volt.reached
 =====

date 99-02-03 time 13:05
 GET pH 12 # 19
 .10ml/div dPH=1/div
 start V .000 ml

STD 1118
 date 99-02-03 time 14:36
 GET pH 12 # 21
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 11.22
 V/ml pH
 EP1 .253 9.37
 EP2 .324 6.42
 stop volt.reached
 =====

date 99-02-03 time 14:36
 GET pH 12 # 21
 .10ml/div dPH=1/div
 start V .000 ml

Pwp 81K
 date 99-02-03 time 14:56
 GET pH 12 # 24
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 5.01
 stop volt.reached
 =====

date 99-02-03 time 14:56
 GET pH 12 # 24
 .10ml/div dPH=1/div
 start V .000 ml

3390
 date 99-02-03 time 14:42
 GET pH 12 # 22
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 9.16
 V/ml pH
 EP1 .023 7.26
 stop volt.reached
 =====

date 99-02-03 time 14:48
 GET pH 12 # 22
 .10ml/div dPH=1/div
 start V .000 ml

3390 Dup
 date 99-02-03 time 14:54
 GET pH 12 # 23
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 9.04
 V/ml pH
 EP1 .024 7.02
 stop volt.reached
 =====

date 99-02-03 time 14:54
 GET pH 12 # 23
 .10ml/div dPH=1/div
 start V .000 ml

3395
 date 99-02-03 time 15:08
 GET pH 12 # 25
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 6.96
 V/ml pH
 EP1 .022 5.73
 stop volt.reached
 =====

date 99-02-03 time 15:09
 GET pH 12 # 25
 .10ml/div d_{pH}=1/div
 start V .000 ml

3395 Dup
 date 99-02-03 time 15:14
 GET pH 12 # 26
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 8.77
 V/ml pH
 EP1 .028 6.18
 stop volt.reached
 =====

date 99-02-03 time 15:14
 GET pH 12 # 26
 .10ml/div d_{pH}=1/div
 start V .000 ml

3400
 date 99-02-03 time 15:21
 GET pH 12 # 27
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 9.83
 V/ml pH
 EP1 .066 7.11
 EP2 .093 5.34
 stop volt.reached
 =====

date 99-02-03 time 15:21
 GET pH 12 # 27
 .10ml/div d_{pH}=1/div
 start V .000 ml

3400 Dup
 date 99-02-03 time 15:26
 GET pH 12 # 28
 Id.#1 268.5
 Id.#2 .1934
 pH(init) 9.92
 V/ml pH
 EP1 .019 9.45
 EP2 .051 7.94
 EP3 .087 5.57
 stop volt.reached
 =====

date 99-02-03 time 15:26
 GET pH 12 # 28
 .10ml/div d_{pH}=1/div
 start V .000 ml

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		STD
Type	Sample Size (mL) SS	0.050
STD	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.253
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code		
OH-01	Concentration of Sample Prep (MOLARITY)	9.79E-01
Matrix	Concentration of Sample in PPM	1.66E+04
SOLID		
Sample #		
STD		
Instrument Code		
PH01	Detection Limit (PPM)	
Analyst		
JIS	OH Molarity = ((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/mL = (OH MOLARITY)*(17.0g/mole)*(1000000µg/g)/(1000mL/L)	
Time		
02:45 PM		STD
	Concentration of Sample in PPM	1.66E+04

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		BLNK-PREP
Type	Sample Size (mL) SS	1.000
BLNK-PREP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.008
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.036
OH-01	Concentration of Sample Prep (MOLARITY)	1.55E-03
Matrix	Concentration of Sample in PPM	5.22E+03
SOLID		
Sample #		
BLNK-PREP	Detection Limit= $((125\mu\text{g}/\text{SS}) \times \text{DF}) \times 1000 / \text{DDF}$	
Instrument Code		
PH01	Detection Limit (PPM)	2.48E+04
Analyst		
JIS	OH Molarity = $((\text{mL HNO}_3) \times (\text{M HNO}_3)) / \text{Sample Size in mL} \times \text{Dilution Factor}$	
Date		
02/03/99	OH in $\mu\text{g/g}$ = $(\text{OH MOLARITY}) \times (17\text{g/mole}) \times (1000000 \mu\text{g/g}) / \text{DDF}$	
Time		
02:45 PM		
		BLNK-PREP
	Concentration of Sample in PPM	<24821
	The Result is < Detection Limit	

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		SAMPLE
Type	Sample Size (mL) SS	1.000
SAMPLE	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.023
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.036
OH-01	Concentration of Sample Prep (MOLARITY)	4.45E-03
Matrix	Concentration of Sample in PPM	1.50E+04
SOLID		
Sample #		
S98T003390	Detection Limit= $((125\mu\text{g} / \text{SS}) * \text{DF}) * 1000 / \text{DDF}$	
Instrument Code		
PH01	Detection Limit (PPM)	2.48E+04
Analyst		
JIS	OH Molarity = $((\text{mL HNO}_3) * (\text{M HNO}_3)) / \text{Sample Size in mL} * \text{Dilution Factor}$	
Date		
02/03/99	OH in $\mu\text{g/g}$ = $(\text{OH MOLARITY}) * (17\text{g/mole}) * (1000000 \mu\text{g/g}) / \text{DDF}$	
Time		
02:45 PM		
		SAMPLE
	Concentration of Sample in PPM	<24821
	The Result is < Detection Limit	

OH (AUTO) : LA-211-102 (D-1)

		DUP
Type	Sample Size (mL) SS	1.000
DUP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.024
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.133
OH-01	Concentration of Sample Prep (MOLARITY)	4.64E-03
Matrix	Concentration of Sample in PPM	1.54E+04
SOLID		
Sample #		
S98T003390	Detection Limit=((125µg /SS)*DF)*1000/DDF	
Instrument Code		
PH01	Detection Limit (PPM)	2.44E+04
Analyst		
JIS	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/g =(OH MOLARITY)*(17g/mole) *(1000000 µg/g)/DDF	
Time		
02:45 PM		
		DUP
Concentration of Sample in PPM		<24352
The Result is < Detection Limit		

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		SAMPLE
Type	Sample Size (mL) SS	1.000
SAMPLE	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.022
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	4.942
OH-01	Concentration of Sample Prep (MOLARITY)	4.25E-03
Matrix	Concentration of Sample in PPM	1.46E+04
SOLID		
Sample #		
S98T003395	Detection Limit=((125µg /SS)*DF)*1000/DDF	
Instrument Code		
PH01	Detection Limit (PPM)	2.53E+04
Analyst		
JIS	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/g =(OH MOLARITY)*(17g/mole) *(1000000 µg/g)/DDF	
Time		
02:45 PM		
		SAMPLE
	Concentration of Sample in PPM	<25293
	The Result is < Detection Limit	

OH (AUTO) : LA-211-102 (D-1)

		DUP
Type	Sample Size (mL) SS	1.000
DUP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.028
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.158
OH-01	Concentration of Sample Prep (MOLARITY)	5.42E-03
Matrix	Concentration of Sample in PPM	1.78E+04
SOLID		
Sample #		
S98T003395	Detection Limit=((125µg /SS)*DF)*1000/DDF	
Instrument Code		
PH01	Detection Limit (PPM)	2.42E+04
Analyst		
JIS	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/g =(OH MOLARITY)*(17g/mole) *(1000000 µg/g)/DDF	
Time		
02:45 PM		
		DUP
	Concentration of Sample in PPM	<24234
The Result is < Detection Limit		

OH (AUTO) : LA-211-102 (D-1)

OH (AUTO) : LA-211-102 (D-1)		SAMPLE
Type	Sample Size (mL) SS	1.000
SAMPLE	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.066
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.068
OH-01	Concentration of Sample Prep (MOLARITY)	1.28E-02
Matrix	Concentration of Sample in PPM	4.28E+04
SOLID		
Sample #		
S98T003400	Detection Limit=((125µg /SS)*DF)*1000/DDF	
Instrument Code		
PH01	Detection Limit (PPM)	2.47E+04
Analyst		
JIS	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/g =(OH MOLARITY)*(17g/mole) *(1000000 µg/g)/DDF	
Time		
02:45 PM		
		SAMPLE
	Concentration of Sample in PPM	4.28E+04

OH (AUTO) : LA-211-102 (D-1)

		DUP
Type	Sample Size (mL) SS	1.000
DUP	Concentration of Titrant (Molarity)	0.1934
Work List	Titrant volume at end-point in mL	0.019
28138	*** Enter Dilution Factor (DF) or 1 ***	1
Test Code	***Enter Digest Factor (DDF) in g/L***	5.072
OH-01	Concentration of Sample Prep (MOLARITY)	3.67E-03
Matrix	Concentration of Sample in PPM	1.23E+04
SOLID		
Sample #		
S98T003400	Detection Limit=((125µg /SS)*DF)*1000/DDF	
Instrument Code		
PH01	Detection Limit (PPM)	2.46E+04
Analyst		
JIS	OH Molarity =((mL HNO3)*(M HNO3))/Sample Size in mL)*Dilution Factor	
Date		
02/03/99	OH in µg/g =(OH MOLARITY)*(17g/mole) *(1000000 µg/g)/DDF	
Time		
02:45 PM		
		DUP
Concentration of Sample in PPM		<24645
The Result is < Detection Limit		

LABCORE Completed Worklist Report for Worklist# 26728

Analyst: slh

Instrument: NH301

Book#: _____

Method: LA-631-001

Rev/Mod C-631-001-3-05-99

Worklist Comment: BY-105 FOR NH3-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	NH3-01	LIQUID	3.98E+02	3.69E+2	92.714 % Recovery	
2 BLNK	0	NH3-01	LIQUID	1	<5.00E+0	ug/mL	
3 SAMPLE	S98T002392 0	NH3-01	LIQUID	6.46E+01	35.000	ug/mL	
4 DUP	S98T002392 0	NH3-01	LIQUID	6.46E+1	1.09E+2	51.152 RPD	
5 TRIPL	S98T002392 0	NH3-01	LIQUID	6.46E+1	<5.50E+1	RPD	
6 SAMPLE	S98T002806 0	NH3-01	LIQUID	N/A	2.76E+02	55.000 ug/mL	
7 DUP	S98T002806 0	NH3-01	LIQUID	2.76E+2	1.78E+2	43.172 RPD	
8 TRIPL	S98T002806 0	NH3-01	LIQUID	2.76E+2	1.89E+2	37.419 RPD	
9 SAMPLE	S98T002807 0	NH3-01	LIQUID	N/A	1.86E+02	55.000 ug/mL	
10 DUP	S98T002807 0	NH3-01	LIQUID	1.86E+2	6.02E+1	102.193 RPD	
11 TRIPL	S98T002807 0	NH3-01	LIQUID	1.86E+2	<5.50E+1	RPD	
12 STD	0	NH3-01	LIQUID	3.98E+02	4.11E+2	103.266 % Recovery	

Final page for worklist# 26728

Analyst Signature _____ Date _____

Mary Tracy 12-29-98
Analyst Signature Date

[Signature] 12/31/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26728

Analyst: SLH Instrument: NH301 Book# SDN19-F

Method: LA-631-001 Rev/Mod C-6 3-25-99

Worklist Comment: BY-105 FOR NH3-01 RTS Prepare 1-10mL dilution. Use 1.5mL

S	Type	Sample#	R	A	Test	Matrix	Group#	Project	Sample Size
1	STD				NH3-01	LIQUID			
2	BLNK				NH3-01	LIQUID			
3	SAMPLE	S98T002392 0			NH3-01	LIQUID	98000436	BY-105 (2)	
		Analytes Requested: NH3-01							
4	DUP	S98T002392 0			NH3-01	LIQUID			
5	SAMPLE	S98T002806 0			NH3-01	LIQUID	98000437	BY-105 (2)	
		Analytes Requested: NH3-01							
6	DUP	S98T002806 0			NH3-01	LIQUID			
7	SAMPLE	S98T002807 0			NH3-01	LIQUID	98000437	BY-105 (2)	
		Analytes Requested: NH3-01							
8	DUP	S98T002807 0			NH3-01	LIQUID			
9	STD				NH3-01	LIQUID			

Final page for worklist # 26728

Andrew Hood Boatright
Analyst Signature Date
12-19-98

Mary Franz 12-29-98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 09:52, 12-18-98

SAMPLE VOL= 25.000 AT 09:52, 12-18-98
ENTERED

EMF= 175.6 mV AT 09:56, 12-18-98

EMF= 174.3 mV AT 09:57, 12-18-98

EMF= 174.0 mV AT 09:57, 12-18-98

EMF= 173.8 mV AT 09:58, 12-18-98

EMF= 173.4 mV AT 09:58, 12-18-98

EMF= 173.3 mV AT 09:58, 12-18-98

EMF= 173.0 mV AT 09:59, 12-18-98

EMF= 172.8 mV AT 09:59, 12-18-98

EMF= 172.7 mV AT 09:59, 12-18-98

EMF= 172.5 mV AT 10:00, 12-18-98

EMF= 172.4 mV AT 10:01, 12-18-98

EMF= 172.4 mV AT 10:02, 12-18-98

EMF= 172.3 mV AT 10:02, 12-18-98

EMF= 172.3 mV AT 10:02, 12-18-98

EMF= 172.3 mV AT 10:03, 12-18-98

EMF= 172.5 mV AT 10:04, 12-18-98

EMF= 172.5 mV AT 10:04, 12-18-98

EMF= 172.8 mV AT 10:05, 12-18-98
ENTERED

STD CONC= 1000 AT 10:05, 12-18-98
ENTERED

STD VOL= .25000 AT 10:05, 12-18-98
ENTERED

EMF= 42.0 mV AT 10:08, 12-18-98

EMF= 41.4 mV AT 10:08, 12-18-98

EMF= 41.4 mV AT 10:09, 12-18-98

EMF= 41.4 mV AT 10:09, 12-18-98

EMF= 41.0 mV AT 10:09, 12-18-98

EMF= 40.9 mV AT 10:10, 12-18-98

EMF= 40.8 mV AT 10:10, 12-18-98
ENTERED

STD VOL= 2.5000 AT 10:11, 12-18-98
ENTERED

EMF=-15.3 mV AT 10:13, 12-18-98

EMF=-15.6 mV AT 10:14, 12-18-98

EMF=-15.8 mV AT 10:14, 12-18-98

EMF=-16.0 mV AT 10:15, 12-18-98

EMF=-16.1 mV AT 10:16, 12-18-98

EMF=-15.9 mV AT 10:17, 12-18-98

EMF=-15.9 mV AT 10:17, 12-18-98
ENTERED

1:NH3 SLOPE=-56.8 mV/DEC
AT 10:17, 12-18-98

1:NH3 CONC= .0471

DOUBLE KNOWN ADDITION SELECTED

BMC

WL 26728

SAMPLE VOL= 25.000 AT 11:03, 12-18-98
ENTERED

EMF= 30.0 mV AT 11:05, 12-18-98

EMF= 29.9 mV AT 11:06, 12-18-98

EMF= 29.9 mV AT 11:06, 12-18-98

EMF= 34.8 mV AT 11:07, 12-18-98

EMF= 34.3 mV AT 11:07, 12-18-98

EMF= 32.2 mV AT 11:08, 12-18-98

EMF= 32.0 mV AT 11:09, 12-18-98

EMF= 31.9 mV AT 11:09, 12-18-98

EMF= 32.0 mV AT 11:09, 12-18-98

EMF= 32.0 mV AT 11:10, 12-18-98

EMF= 32.1 mV AT 11:10, 12-18-98

EMF= 31.9 mV AT 11:11, 12-18-98

EMF= 31.9 mV AT 11:11, 12-18-98

EMF= 32.3 mV AT 11:12, 12-18-98

EMF= 32.2 mV AT 11:12, 12-18-98

EMF= 32.2 mV AT 11:12, 12-18-98

EMF= 31.9 mV AT 11:13, 12-18-98

EMF= 32.0 mV AT 11:13, 12-18-98

EMF= 32.0 mV AT 11:13, 12-18-98

EMF= 31.7 mV AT 11:14, 12-18-98

EMF= 32.1 mV AT 11:14, 12-18-98

EMF= 32.6 mV AT 11:15, 12-18-98

EMF= 32.4 mV AT 11:16, 12-18-98

EMF= 32.4 mV AT 11:16, 12-18-98

EMF= 31.7 mV AT 11:16, 12-18-98

ENTERED

STD CONC= 1000 AT 11:17, 12-18-98
ENTERED

STD VOL= .25000 AT 11:17, 12-18-98
ENTERED

EMF= 22.5 mV AT 11:18, 12-18-98

EMF= 20.9 mV AT 11:19, 12-18-98

EMF= 21.0 mV AT 11:20, 12-18-98

EMF= 21.5 mV AT 11:20, 12-18-98

EMF= 21.5 mV AT 11:21, 12-18-98

EMF= 18.8 mV AT 11:23, 12-18-98

EMF= 18.7 mV AT 11:23, 12-18-98

EMF= 18.6 mV AT 11:23, 12-18-98

EMF= 18.9 mV AT 11:24, 12-18-98
ENTERED

STD VOL= 2.5000 AT 11:24, 12-18-98
ENTERED

EMF=-19.4 mV AT 11:27, 12-18-98

EMF=-19.6 mV AT 11:27, 12-18-98

EMF=-20.0 mV AT 11:29, 12-18-98

EMF=-19.8 mV AT 11:30, 12-18-98

EMF=-19.7 mV AT 11:31, 12-18-98

EMF=-19.6 mV AT 11:31, 12-18-98
ENTERED

1:NH3 SLOPE=-58.3 mV/DEC
AT 11:31, 12-18-98

1:NH3 CONCEN= 14.8

STD 1ml 50M19-F

DOUBLE KNOWN ADDITION SELECTED
AT 12:56, 12-18-98

SAMPLE VOL= 25.000 AT 12:56, 12-18-98
ENTERED

EMF= 101.1 mV AT 13:03, 12-18-98

EMF= 100.6 mV AT 13:04, 12-18-98

EMF= 100.2 mV AT 13:04, 12-18-98

EMF= 100.0 mV AT 13:04, 12-18-98

EMF= 99.8 mV AT 13:05, 12-18-98

EMF= 99.6 mV AT 13:05, 12-18-98

EMF= 99.5 mV AT 13:05, 12-18-98

EMF= 99.3 mV AT 13:06, 12-18-98

EMF= 99.1 mV AT 13:06, 12-18-98
ENTERED

STD CONC= 1000 AT 13:08, 12-18-98
ENTERED

STD VOL= .25000 AT 13:08, 12-18-98
ENTERED

EMF= 43.5 mV AT 13:11, 12-18-98

EMF= 41.2 mV AT 13:12, 12-18-98

EMF= 40.9 mV AT 13:13, 12-18-98

EMF= 40.6 mV AT 13:13, 12-18-98

EMF= 40.3 mV AT 13:13, 12-18-98

EMF= 40.2 mV AT 13:14, 12-18-98

EMF= 40.2 mV AT 13:14, 12-18-98
ENTERED

STD VOL= 2.5000 AT 13:15, 12-18-98
ENTERED

EMF=-14.0 mV AT 13:19, 12-18-98

EMF=-14.5 mV AT 13:19, 12-18-98

EMF=-14.8 mV AT 13:19, 12-18-98

EMF=-14.9 mV AT 13:20, 12-18-98

EMF=-15.1 mV AT 13:20, 12-18-98

EMF=-15.1 mV AT 13:20, 12-18-98

EMF=-15.3 mV AT 13:21, 12-18-98

EMF=-15.4 mV AT 13:22, 12-18-98

EMF=-15.4 mV AT 13:22, 12-18-98
ENTERED

1:NH3 SLOPE=-57.8 mV/DEC
AT 13:22, 12-18-98

1:NH3 CONC= 1.05

598T002806

1/-10ml dil 1ml ss sample

DOUBLE KNOWN ADDITION SELECTED
AT 13:25, 12-18-98

SAMPLE VOL= 25.000 AT 13:25, 12-18-98
ENTERED

EMF= 105.4 mV AT 13:35, 12-18-98

EMF= 105.9 mV AT 13:36, 12-18-98

EMF= 106.4 mV AT 13:36, 12-18-98

EMF= 106.7 mV AT 13:37, 12-18-98

EMF= 107.1 mV AT 13:37, 12-18-98

EMF= 107.3 mV AT 13:37, 12-18-98
ENTERED

STD CONC= 1000 AT 13:38, 12-18-98
ENTERED

STD VOL= .25000 AT 13:38, 12-18-98
ENTERED

EMF= 41.7 mV AT 13:41, 12-18-98

EMF= 41.1 mV AT 13:41, 12-18-98

EMF= 40.8 mV AT 13:42, 12-18-98

EMF= 40.5 mV AT 13:42, 12-18-98

EMF= 40.3 mV AT 13:43, 12-18-98

EMF= 40.3 mV AT 13:43, 12-18-98
ENTERED

STD VOL= 2.5000 AT 13:43, 12-18-98
ENTERED

EMF=-14.3 mV AT 13:46, 12-18-98

EMF=-14.6 mV AT 13:47, 12-18-98

EMF=-14.9 mV AT 13:47, 12-18-98

EMF=-15.1 mV AT 13:49, 12-18-98

EMF=-15.0 mV AT 13:50, 12-18-98

EMF=-14.9 mV AT 13:50, 12-18-98
ENTERED

1:NH3 SLOPE=-56.7 mV/DEC
AT 13:50, 12-18-98

1:NH3 CONC= .696

398T002806

1-10ml del 1ml ss deep

DOUBLE KNOWN ADDITION SELECTED
AT 13:53, 12-18-98

SAMPLE VOL= 25.000 AT 13:53, 12-18-98
ENTERED

EMF= 100.4 mV AT 14:01, 12-18-98

EMF= 103.2 mV AT 14:03, 12-18-98

EMF= 103.9 mV AT 14:04, 12-18-98

EMF= 104.3 mV AT 14:04, 12-18-98

EMF= 104.6 mV AT 14:05, 12-18-98

EMF= 105.0 mV AT 14:05, 12-18-98

EMF= 105.2 mV AT 14:05, 12-18-98
ENTERED

STD CONC= 1000 AT 14:06, 12-18-98
ENTERED

HNF-1663 REV. 0

STD VOL= .25000 AT 14:06, 12-18-98
ENTERED

EMF= 41.5 mV AT 14:09, 12-18-98

EMF= 41.0 mV AT 14:10, 12-18-98

EMF= 40.8 mV AT 14:10, 12-18-98

EMF= 40.3 mV AT 14:12, 12-18-98

EMF= 40.3 mV AT 14:13, 12-18-98

EMF= 40.4 mV AT 14:13, 12-18-98

EMF= 40.4 mV AT 14:14, 12-18-98

EMF= 40.6 mV AT 14:15, 12-18-98

EMF= 40.6 mV AT 14:16, 12-18-98

EMF= 40.6 mV AT 14:16, 12-18-98
ENTERED

STD VOL= 2.5000 AT 14:18, 12-18-98
ENTERED

EMF=-13.4 mV AT 14:21, 12-18-98

EMF=-14.2 mV AT 14:21, 12-18-98

EMF=-14.5 mV AT 14:22, 12-18-98

EMF=-14.9 mV AT 14:23, 12-18-98

EMF=-14.9 mV AT 14:24, 12-18-98

EMF=-14.9 mV AT 14:24, 12-18-98

EMF=-14.8 mV AT 14:26, 12-18-98

EMF=-14.7 mV AT 14:26, 12-18-98

EMF=-14.6 mV AT 14:27, 12-18-98

EMF=-14.6 mV AT 14:28, 12-18-98

EMF=-14.3 mV AT 14:29, 12-18-98

EMF=-14.2 mV AT 14:30, 12-18-98

EMF=-14.1 mV AT 14:30, 12-18-98

EMF=-13.9 mV AT 14:31, 12-18-98

EMF=-13.7 mV AT 14:32, 12-18-98

EMF=-13.6 mV AT 14:32, 12-18-98

EMF=-13.3 mV AT 14:34, 12-18-98

EMF=-13.4 mV AT 14:34, 12-18-98

EMF=-13.5 mV AT 14:34, 12-18-98

ENTERED

1:NH3 SLOPE=-55.6 mV/DEC

AT 14:34, 12-18-98

1:NH3 CONC= .733

598700 2806 Trp 1-10ml del /ml ss

DOUBLE KNOWN ADDITION SELECTED

AT 15:13, 12-18-98

SAMPLE VOL= 25.000 AT 15:13, 12-18-98

ENTERED

EMF= 108.2 mV AT 15:22, 12-18-98

EMF= 107.6 mV AT 15:22, 12-18-98

EMF= 107.3 mV AT 15:23, 12-18-98

EMF= 106.9 mV AT 15:23, 12-18-98

EMF= 106.7 mV AT 15:24, 12-18-98

EMF= 106.6 mV AT 15:24, 12-18-98

ENTERED

STD CONC= 1000 AT 15:24, 12-18-98

ENTERED

STD VOL= .25000 AT 15:24, 12-18-98

ENTERED

EMF= 41.6 mV AT 15:29, 12-18-98

EMF= 41.2 mV AT 15:29, 12-18-98

EMF= 40.9 mV AT 15:30, 12-18-98

EMF= 40.6 mV AT 15:30, 12-18-98

EMF= 40.5 mV AT 15:31, 12-18-98

EMF= 40.4 mV AT 15:31, 12-18-98

EMF= 40.3 mV AT 15:32, 12-18-98

EMF= 40.3 mV AT 15:32, 12-18-98

ENTERED

STD VOL= 2.5000 AT 15:32, 12-18-98

ENTERED

EMF=-14.1 mV AT 15:37, 12-18-98

EMF=-14.4 mV AT 15:37, 12-18-98

EMF=-14.6 mV AT 15:37, 12-18-98

EMF=-15.0 mV AT 15:38, 12-18-98

EMF=-15.0 mV AT 15:39, 12-18-98

EMF=-15.0 mV AT 15:39, 12-18-98
ENTERED

1:NH3 SLOPE=-56.8 mV/DEC
AT 15:39, 12-18-98

1:NH3 CONC= .724

5987002807 sam 1-10 ml del /ml 23

DOUBLE KNOWN ADDITION SELECTED
AT 15:41, 12-18-98

SAMPLE VOL= 25.000 AT 15:41, 12-18-98
ENTERED

EMF= 121.9 mV AT 15:55, 12-18-98

EMF= 126.1 mV AT 15:56, 12-18-98

EMF= 128.1 mV AT 15:56, 12-18-98

EMF= 130.8 mV AT 15:57, 12-18-98

EMF= 132.6 mV AT 15:59, 12-18-98

EMF= 133.5 mV AT 15:59, 12-18-98

EMF= 133.7 mV AT 15:59, 12-18-98
ENTERED

STD CONC= 1000 AT 15:59, 12-18-98
ENTERED

STD VOL= .25000 AT 15:59, 12-18-98
ENTERED

EMF= 41.8 mV AT 16:04, 12-18-98

EMF= 41.6 mV AT 16:04, 12-18-98

EMF= 41.4 mV AT 16:05, 12-18-98

EMF= 41.4 mV AT 16:05, 12-18-98

EMF= 41.2 mV AT 16:06, 12-18-98

EMF= 41.1 mV AT 16:06, 12-18-98

EMF= 41.1 mV AT 16:07, 12-18-98

EMF= 41.2 mV AT 16:07, 12-18-98

EMF= 41.2 mV AT 16:08, 12-18-98

EMF= 41.3 mV AT 16:09, 12-18-98

EMF= 41.4 mV AT 16:09, 12-18-98

EMF= 41.7 mV AT 16:11, 12-18-98

~~EMF= 41.8 mV AT 16:11, 12-18-98~~

EMF= 41.9 mV AT 16:12, 12-18-98

EMF= 42.3 mV AT 16:12, 12-18-98
ENTERED

STD VOL= 2.5000 AT 16:12, 12-18-98
ENTERED

EMF=-14.0 mV AT 16:16, 12-18-98

EMF=-14.3 mV AT 16:17, 12-18-98

EMF=-14.9 mV AT 16:18, 12-18-98

EMF=-14.9 mV AT 16:18, 12-18-98

EMF=-14.9 mV AT 16:18, 12-18-98

EMF=-14.9 mV AT 16:19, 12-18-98
ENTERED

1:NH3 SLOPE=-57.8 mV/DEC
AT 16:19, 12-18-98

1:NH3 CONCEN= .266

S98T002807 dup 6-10ml del 1ml SS

DOUBLE KNOWN ADDITION SELECTED
AT 16:21, 12-18-98

SAMPLE VOL= 25.000 AT 16:21, 12-18-98
ENTERED

EMF= 139.2 mV AT 16:36, 12-18-98

EMF= 145.3 mV AT 16:40, 12-18-98

EMF= 146.8 mV AT 16:41, 12-18-98

EMF= 147.1 mV AT 16:42, 12-18-98

EMF= 147.5 mV AT 16:42, 12-18-98

EMF= 147.7 mV AT 16:42, 12-18-98
ENTERED

STD CONC= 1000 AT 16:42, 12-18-98
ENTERED

STD VOL= .25000 AT 16:43, 12-18-98
ENTERED

EMF= 44.5 mV AT 16:46, 12-18-98

EMF= 44.1 mV AT 16:46, 12-18-98

EMF= 43.7 mV AT 16:47, 12-18-98

EMF= 43.4 mV AT 16:47, 12-18-98

EMF= 43.3 mV AT 16:48, 12-18-98

EMF= 43.1 mV AT 16:48, 12-18-98

EMF= 43.0 mV AT 16:49, 12-18-98

EMF= 43.1 mV AT 16:49, 12-18-98
ENTERED

STD VOL= 2.5000 AT 16:49, 12-18-98
ENTERED

EMF=-13.1 mV AT 16:53, 12-18-98

EMF=-13.4 mV AT 16:53, 12-18-98

EMF=-13.9 mV AT 16:54, 12-18-98

EMF=-13.7 mV AT 16:55, 12-18-98

EMF=-13.7 mV AT 16:55, 12-18-98

EMF=-13.5 mV AT 16:55, 12-18-98
ENTERED

1:NH3 SLOPE=-56.9 mV/DEC
AT 16:55, 12-18-98

1:NH3 CONC= .146

598700 2807 Trp 1-10/4ml 26

DOUBLE KNOWN ADDITION SELECTED
AT 16:57, 12-18-98

SAMPLE VOL= 25.000 AT 16:57, 12-18-98
ENTERED

EMF= 126.5 mV AT 17:13, 12-18-98

EMF= 127.2 mV AT 17:14, 12-18-98

EMF= 128.5 mV AT 17:14, 12-18-98

EMF= 129.5 mV AT 17:15, 12-18-98

EMF= 130.3 mV AT 17:16, 12-18-98

EMF= 130.5 mV AT 17:16, 12-18-98
ENTERED

STD CONC= 1000 AT 17:16, 12-18-98
ENTERED

STD VOL= .25000 AT 17:16, 12-18-98
ENTERED

EMF= 44.7 mV AT 17:20, 12-18-98

EMF= 44.3 mV AT 17:21, 12-18-98

EMF= 43.5 mV AT 17:22, 12-18-98

EMF= 43.1 mV AT 17:22, 12-18-98

EMF= 42.5 mV AT 17:23, 12-18-98
ENTERED

STD VOL= 2.5000 AT 17:23, 12-18-98
ENTERED

EMF=-12.5 mV AT 17:28, 12-18-98

EMF=-12.8 mV AT 17:28, 12-18-98

EMF=-13.0 mV AT 17:28, 12-18-98

EMF=-13.2 mV AT 17:29, 12-18-98

EMF=-13.4 mV AT 17:29, 12-18-98

EMF=-13.6 mV AT 17:30, 12-18-98

EMF=-13.4 mV AT 17:30, 12-18-98
ENTERED

MODEL KNOWN ADDITION SELECTED
AT 23:26, 12-18-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 23:26, 12-18-98
ENTERED

EMF= 32.3 mV AT 23:30, 12-18-98

EMF= 32.3 mV AT 23:30, 12-18-98
ENTERED

STD CONC= 1000 AT 23:30, 12-18-98
ENTERED

STD VOL= .25000 AT 23:30, 12-18-98
ENTERED

EMF= 20.7 mV AT 23:33, 12-18-98

EMF= 20.8 mV AT 23:33, 12-18-98
ENTERED

STD VOL= 2.5000 AT 23:34, 12-18-98
ENTERED

EMF=-15.6 mV AT 23:38, 12-18-98

EMF=-15.6 mV AT 23:38, 12-18-98
ENTERED

1:NH3 SLOPE=-57.1 mV/DEC
AT 23:38, 12-18-98

1:NH3 CONC= 16.5

STD End

SON19-F

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-¹⁴D) ³²⁵⁴⁹

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 14.800
STD	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
LIQUID		
Batch Number		
98005334		
Retun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.69E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
SLH		
Date Complete		
12/29/98		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/18/98	QC FOUND (µg)	3.69E+02
Analysis Time		
07:00 PM		
Sample Point		
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist: <i>N/A</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: BLANK2

AMMONIA (NH3) : LA-631-001 (C-0) ^{1 MD} ₃₋₂₆₋₉₉

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.047
Work (List)	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.00E+00
Sample #		
BLNK	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 5.00E+00
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM6

AMMONIA (NH3) : LA-631-001 (C-9)¹⁰⁰⁰₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 1.050
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Retun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.76E+02
Sample #		
S98T002806	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	2.76E+02
BY105		

Analyst:	SLH	Date: 12/29/98
Signature of Chemist:	<i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP7

AMMONIA (NH3) : LA-631-001 (C-0) ^{1.40} _{3 25-99}

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.696
DUP	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.78E+02
Sample #		
S98T002806	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.78E+02
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist: <i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: TRIPL8

AMMONIA (NH3) : LA-631-001 (C-0) ^{1.00} 3-25-91

LIQUIDS/SOLIDS

		TRIPL
Type	Instrument Data (µg/mL)	ID 0.733
TRIPL	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.89E+02
Sample #		
S98T002806	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.89E+02
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist: <i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

807

WORKBOOK PAGE: SAM9

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁰⁰ 325-11

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.724
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.86E+02
Sample #		
S98T002807	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.86E+02
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP10

AMMONIA (NH3) : LA-631-001 (C-0) ¹3-25-99

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.266
DUP	Blank Result from the Instrument (µg/mL)	BR 0.047
WorkList	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 6.02E+01
Sample #		
S98T002807	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	6.02E+01
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: TRIPL11

AMMONIA (NH₃) : LA-631-001 (C-0) ¹₃₂₅₄₉

LIQUIDS/SOLIDS

		TRIPL
Type	Instrument Data (µg/mL)	ID 0.146
TRIPL	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005334		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.50E+01
Sample #		
S98T002807	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
12/29/98		
Analysis Date		
12/18/98		
Analysis Time		
07:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 5.50E+01
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: ST_END12

AMMONIA (NH3) : LA-631-001 (C-0) 3-25-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 16.500
STD	Blank Result from the Instrument (µg/mL)	BR 0.047
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26728	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
LIQUID		
Batch Number		
98005334		
Re run		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.11E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
SLH		
Date Complete		
12/29/98		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/18/98	QC FOUND (µg)	4.11E+02
Analysis Time		
07:00 PM		
Sample Point		
BY105		

Analyst:	SLH Date: 12/29/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

LABCORE Completed Worklist Report for Worklist# 26730

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 JMV

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD	0		NH3-01	LIQUID	4.25E+02	4.56E+2	107.294 %	Recovery
2	BLNK	0		NH3-01	LIQUID	1	<1.00E+1		ug/mL
3	SAMPLE	S98T002955	0	NH3-01	LIQUID	N/A	4.77E+01	10.000	ug/mL
4	DUP	S98T002955	0	NH3-01	LIQUID	4.77E+1	4.18E+1	13.184	RPD
5	SPK	S98T002955	0	NH3-01	LIQUID	1.06E+02	1.12E+02	105.660 %	Recovery
6	STD	0		NH3-01	LIQUID	4.25E+02	4.49E+2	105.647 %	Recovery

Final page for worklist# 26730

11/16/98
_____
Analyst Signature Date

_____
Analyst Signature Date

11/17/98
_____
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26730

Analyst: MF Instrument: NH301 Book# 50 N19-P

Method: LA-631-001 Rev/Mod C-0102-05-11

Worklist Comment: BY-105 FOR NH3-01 JMV

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			NH3-01	LIQUID		
2 BLNK			NH3-01	LIQUID		
3 SAMPLE	S98T002955 0		NH3-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T002955 0		NH3-01	LIQUID		
5 SPK	S98T002955 0		NH3-01	LIQUID		
6 STD			NH3-01	LIQUID		

Final page for worklist # 26730

Mary Tracy 10-27-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

SAMPLE VOL= 25.000 AT 10:25, 10-27-98
ENTERED

HNF-1663 REV. 0

MF= 45.1 mV AT 10:25, 10-27-98

MF= 45.1 mV AT 10:25, 10-27-98

MF= 45.3 mV AT 10:26, 10-27-98

MF= 45.4 mV AT 10:26, 10-27-98

MF= 45.9 mV AT 10:27, 10-27-98

MF= 46.1 mV AT 10:28, 10-27-98

MF= 46.3 mV AT 10:28, 10-27-98

MF= 46.5 mV AT 10:29, 10-27-98

MF= 46.5 mV AT 10:29, 10-27-98

MF= 46.8 mV AT 10:30, 10-27-98

MF= 46.9 mV AT 10:30, 10-27-98

MF= 47.1 mV AT 10:31, 10-27-98

EMF= 47.1 mV AT 10:31, 10-27-98

ENTERED

STD CONC= 1000 AT 10:31, 10-27-98

ENTERED

STD VOL= .25000 AT 10:31, 10-27-98

ENTERED

EMF= 28.6 mV AT 10:33, 10-27-98

EMF= 28.3 mV AT 10:34, 10-27-98

EMF= 28.2 mV AT 10:34, 10-27-98

EMF= 28.2 mV AT 10:35, 10-27-98

EMF= 28.2 mV AT 10:35, 10-27-98

EMF= 28.3 mV AT 10:36, 10-27-98

ENTERED

STD VOL= 2.5000 AT 10:36, 10-27-98

ENTERED

EMF=-16.4 mV AT 10:38, 10-27-98

EMF=-16.7 mV AT 10:39, 10-27-98

EMF=-16.7 mV AT 10:41, 10-27-98

ENTERED

1:NH3 SLOPE=-59.9 mV/DEC

AT 10:41, 10-27-98

1:NH3 CONC= 9.26

.500 ml 50N19-D

814

DOUBLE KNOWN ADDITION SELECTED
AT 14:18, 10-27-98

AMPLE VOL= 25.000 AT 14:19, 10-27-98
ENTERED

MF= 149.2 mV AT 14:28, 10-27-98

MF= 150.5 mV AT 14:29, 10-27-98

MF= 151.1 mV AT 14:30, 10-27-98

MF= 152.0 mV AT 14:30, 10-27-98

MF= 152.1 mV AT 14:31, 10-27-98
ENTERED

TD CONC= 1000 AT 14:31, 10-27-98
ENTERED

TD VOL= .25000 AT 14:31, 10-27-98
ENTERED

MF= 42.4 mV AT 14:34, 10-27-98

MF= 42.2 mV AT 14:34, 10-27-98

MF= 42.2 mV AT 14:34, 10-27-98
ENTERED

STD VOL= 2.5000 AT 14:35, 10-27-98
ENTERED

MF=-16.2 mV AT 14:37, 10-27-98

MF=-16.4 mV AT 14:37, 10-27-98

MF=-16.4 mV AT 14:37, 10-27-98
ENTERED

NH3 SLOPE=-58.9 mV/DEC
AT 14:37, 10-27-98

NH3 CONC= .137

BLK

DOUBLE KNOWN ADDITION SELECTED
AT 14:42, 10-27-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 14:43, 10-27-98
ENTERED

MF= 101.2 mV AT 14:46, 10-27-98

MF= 101.7 mV AT 14:47, 10-27-98

MF= 102.1 mV AT 14:47, 10-27-98

MF= 102.3 mV AT 14:47, 10-27-98
ENTERED

TD CONC= 1000 AT 14:48, 10-27-98
ENTERED

TD VOL= .25000 AT 14:48, 10-27-98
ENTERED

MF= 42.0 mV AT 14:50, 10-27-98

MF= 41.7 mV AT 14:51, 10-27-98

MF= 41.6 mV AT 14:51, 10-27-98
ENTERED

TD VOL= 2.5000 AT 14:51, 10-27-98
ENTERED

MF=-15.9 mV AT 14:53, 10-27-98

MF=-16.4 mV AT 14:54, 10-27-98

MF=-16.5 mV AT 14:54, 10-27-98
ENTERED

:NH3 SLOPE=-60.5 mV/DEC
AT 14:54, 10-27-98

-NH3 CONC= 1.09

500ml 598T2955

DOUBLE KNOWN ADDITION SELECTED
AT 14:56, 10-27-98

HNF-1663 REV. 0

AMPLE VOL= 25.000 AT 14:57, 10-27-98
ENTERED

MF= 101.7 mV AT 15:02, 10-27-98

MF= 102.3 mV AT 15:03, 10-27-98

MF= 103.0 mV AT 15:04, 10-27-98

MF= 103.2 mV AT 15:04, 10-27-98
ENTERED

TD CONC= 1000 AT 15:04, 10-27-98
ENTERED

TD VOL= .25000 AT 15:04, 10-27-98
ENTERED

MF= 41.5 mV AT 15:07, 10-27-98

MF= 41.2 mV AT 15:07, 10-27-98

MF= 41.1 mV AT 15:08, 10-27-98
ENTERED

TD VOL= 2.5000 AT 15:08, 10-27-98
ENTERED

MF=-15.5 mV AT 15:10, 10-27-98

MF=-15.9 mV AT 15:11, 10-27-98

MF=-16.0 mV AT 15:11, 10-27-98

MF=-16.0 mV AT 15:12, 10-27-98
ENTERED

:NH3 SLOPE=-59.2 mV/DEC
AT 15:12, 10-27-98

:NH3 CONC= .972

.500 ml 2955 Dup

DOUBLE KNOWN ADDITION SELECTED
AT 15:14, 10-27-98

AMPLE VOL= 25.000 AT 15:15, 10-27-98
ENTERED

EMF= 56.4 mV AT 15:15, 10-27-98

EMF= 58.1 mV AT 15:16, 10-27-98

EMF= 58.8 mV AT 15:17, 10-27-98

EMF= 59.2 mV AT 15:17, 10-27-98

EMF= 59.5 mV AT 15:17, 10-27-98

EMF= 59.9 mV AT 15:18, 10-27-98

EMF= 60.0 mV AT 15:18, 10-27-98
ENTERED

STD CONC= 1000 AT 15:19, 10-27-98
ENTERED

STD VOL= .25000 AT 15:19, 10-27-98
ENTERED

EMF= 33.9 mV AT 15:22, 10-27-98

EMF= 33.7 mV AT 15:22, 10-27-98

EMF= 33.7 mV AT 15:23, 10-27-98
ENTERED

STD VOL= 2.5000 AT 15:24, 10-27-98
ENTERED

EMF=-15.4 mV AT 15:25, 10-27-98

EMF=-15.6 mV AT 15:26, 10-27-98

EMF=-15.7 mV AT 15:26, 10-27-98

EMF=-15.8 mV AT 15:26, 10-27-98

EMF=-15.7 mV AT 15:27, 10-27-98

EMF=-15.7 mV AT 15:27, 10-27-98
ENTERED

..NH3 SLOPE=-59.6 mV/DEC
AT 15:27, 10-27-98

..NH3 CONC= 5.58

.500ml 2955 Spk .250 50N19-D

DOUBLE KNOWN ADDITION SELECTED
AT 15:29, 10-27-98

SAMPLE VOL= 25.000 AT 15:29, 10-27-98
ENTERED

EMF= 45.6 mV AT 15:30, 10-27-98

EMF= 45.9 mV AT 15:31, 10-27-98

EMF= 46.2 mV AT 15:31, 10-27-98

EMF= 46.5 mV AT 15:32, 10-27-98

EMF= 46.7 mV AT 15:32, 10-27-98

EMF= 46.7 mV AT 15:33, 10-27-98

EMF= 46.9 mV AT 15:33, 10-27-98

EMF= 47.0 mV AT 15:33, 10-27-98
ENTERED

STD CONC= 1000 AT 15:34, 10-27-98
ENTERED

STD VOL= .25000 AT 15:34, 10-27-98
ENTERED

EMF= 28.4 mV AT 15:36, 10-27-98

EMF= 28.2 mV AT 15:36, 10-27-98

EMF= 28.1 mV AT 15:37, 10-27-98

EMF= 28.1 mV AT 15:37, 10-27-98
ENTERED

STD VOL= 2.5000 AT 15:38, 10-27-98
ENTERED

EMF=-15.7 mV AT 15:39, 10-27-98

EMF=-16.5 mV AT 15:40, 10-27-98

EMF=-16.7 mV AT 15:40, 10-27-98

EMF=-16.8 mV AT 15:41, 10-27-98

EMF=-16.8 mV AT 15:41, 10-27-98
ENTERED

1:NH3 SLOPE=-59.6 mV/DEC
AT 15:41, 10-27-98

1:NH3 CONC= 9.12

. 500ml 50N19-D

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-9)

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 9.260
STD	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005336		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.56E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
10/27/98	QC FOUND (µg)	4.56E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist: <i>N/A</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: BLANK2

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.137
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005336		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
BLK	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/27/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0)¹³₃₅₄₉

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 1.090
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005336		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.77E+01
Sample #		
S98T002955	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/27/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	4.77E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) 8-25-99

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.972
DUP	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005336		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.18E+01
Sample #		
S98T002955	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/27/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	4.18E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (C-0) ^{1 mo} 2-2-98

LIQUIDS/SOLIDS

		SPK
Type	Instrument Data (µg/mL)	ID 5.580
SPK	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	Spike Book Number	SPK 50N19D
NH3-01	Spike Value (µg/mL)	SPK VAL 4.25E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK 0.250
LIQUID		
Batch Number		
98005336		
Rerun	Sample Instrument Data (µg/mL)	SAM ID 1.09
0	Sample Volume of Sample (mL)	SAM VSAM 0.500
Sample Prep	Sample Final Volume (mL)	SAM FVOL 25.0
N/A		
Sample #		
S98T002955		
Instrument Code		
NH301		
Prepared By		
JDS		
Chemist		
MJL	QC ACTUAL = SPK VAL * VOL SPK	
Analyst	QC FOUND = ((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	1.06E+02
10/27/98	QC FOUND (µg)	1.12E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF	Date: 11/10/98
Signature of Chemist:	<i>MF</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END6

AMMONIA (NH3) : LA-631-001 (C-~~6~~¹)₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 9.120
STD	Blank Result from the Instrument (µg/mL)	BR 0.137
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26730	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005336		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.49E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
10/27/98	QC FOUND (µg)	4.49E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 26842

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	BLNK	0		NH3-01	LIQUID	1	<5.00E+0		ug/mL
2	STD	0		NH3-01	LIQUID	3.98E+02	4.67E+2	117.337 % Recovery	
3	SAMPLE	S98T003002	0	NH3-01	LIQUID	N/A	6.49E+02	55.000	ug/mL
4	DUP	S98T003002	0	NH3-01	LIQUID	6.49E+2	5.80E+2	11.229 RPD	
5	SAMPLE	S98T003003	0	NH3-01	LIQUID	N/A	<5.50E+01	55.000	ug/mL
6	DUP	S98T003003	0	NH3-01	LIQUID	<5.50E+1	5.54E+1	RPD	
7	STD	0		NH3-01	LIQUID	3.98E+02	5.02E+2	126.131 % Recovery	

Final page for worklist# 26842

Analyst Signature _____ Date _____

Mary J. May 12-30-98
Analyst Signature *J* Date

MLA 12/31/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26842

Analyst: MF Instrument: NH301 Book# 50N19F

Method: LA-631-001 Rev/Mod C-10-13-98

Worklist Comment: BY-105 FOR NH3-01 RTS

Performa 1-10 dilution. Use

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	BLNK				NH3-01	LIQUID		
2	STD				NH3-01	LIQUID		
3	SAMPLE	S98T003002 0			NH3-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: NH3-01						
4	DUP	S98T003002 0			NH3-01	LIQUID		
5	SAMPLE	S98T003003 0			NH3-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: NH3-01						
6	DUP	S98T003003 0			NH3-01	LIQUID		
7	STD				NH3-01	LIQUID		

1.5ml sample size.
12/14/98

Final page for worklist # 26842

Mary Fran 12-16-98
Analyst Signature Date

Mary Fran 12-30-98
Analyst Signature Date

Data Entry Comments:

Used 1-10 dil w/1ml sample size

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED

AT 08:56, 12-16-98

AMPLE VOL= 25.000 AT 08:56, 12-16-98
ENTERED

MF= 27.7 mV AT 08:59, 12-16-98

MF= 28.0 mV AT 08:59, 12-16-98

MF= 28.2 mV AT 09:00, 12-16-98

MF= 28.4 mV AT 09:00, 12-16-98

MF= 28.4 mV AT 09:00, 12-16-98

MF= 28.6 mV AT 09:01, 12-16-98

MF= 28.8 mV AT 09:01, 12-16-98
ENTEREDTD CONC= 1000 AT 09:01, 12-16-98
ENTEREDTD VOL= .25000 AT 09:02, 12-16-98
ENTERED

MF= 18.5 mV AT 09:04, 12-16-98

MF= 18.1 mV AT 09:04, 12-16-98

MF= 18.1 mV AT 09:04, 12-16-98

MF= 18.0 mV AT 09:05, 12-16-98
ENTEREDTD VOL= 2.5000 AT 09:05, 12-16-98
ENTERED

MF=-17.9 mV AT 09:08, 12-16-98

MF=-18.3 mV AT 09:09, 12-16-98

MF=-18.3 mV AT 09:10, 12-16-98
ENTERED:NH3 SLOPE=-59.4 mV/DEC
AT 09:10, 12-16-98

:NH3 CONC= 18.7 50019-F 1 ml

DOUBLE KNOWN ADDITION SELECTED

HNF-1663 REV. 0

AT 08:31, 12-16-98

SAMPLE VOL= 25.000 AT 08:31, 12-16-98
ENTERED

MF= 196.3 mV AT 08:32, 12-16-98

MF= 193.7 mV AT 08:33, 12-16-98

MF= 193.3 mV AT 08:34, 12-16-98

MF= 192.9 mV AT 08:34, 12-16-98

MF= 192.5 mV AT 08:34, 12-16-98

MF= 192.1 mV AT 08:35, 12-16-98

MF= 191.5 mV AT 08:35, 12-16-98

MF= 191.4 mV AT 08:35, 12-16-98
ENTERED

TD CONC= 1000 AT 08:36, 12-16-98
ENTERED

TD VOL= .25000 AT 08:36, 12-16-98
ENTERED

MF= 50.4 mV AT 08:45, 12-16-98

MF= 49.9 mV AT 08:46, 12-16-98

MF= 49.6 mV AT 08:46, 12-16-98

MF= 49.5 mV AT 08:46, 12-16-98
ENTERED

TD VOL= 2.5000 AT 08:46, 12-16-98
ENTERED

MF=-8.6 mV AT 08:53, 12-16-98

MF=-9.0 mV AT 08:53, 12-16-98

MF=-9.5 mV AT 08:54, 12-16-98

MF=-9.5 mV AT 08:54, 12-16-98
ENTERED

:NH3 SLOPE=-59.1 mV/DEC
AT 08:54, 12-16-98

:NH3 CONC= .0394 *BLK*

DOUBLE KNOWN ADDITION SELECTED
AT 10:35, 12-16-98

HNF-1663 REV. 0

AMPLE VOL= 25.000 AT 10:35, 12-16-98
ENTERED

MF= 81.9 mV AT 10:40, 12-16-98

MF= 81.5 mV AT 10:40, 12-16-98

MF= 81.2 mV AT 10:40, 12-16-98

MF= 81.1 mV AT 10:41, 12-16-98
ENTERED

TD CONC= 1000 AT 10:41, 12-16-98
ENTERED

TD VOL= .25000 AT 10:41, 12-16-98
ENTERED

MF= 40.9 mV AT 10:45, 12-16-98

MF= 40.6 mV AT 10:46, 12-16-98

MF= 40.3 mV AT 10:46, 12-16-98

MF= 40.2 mV AT 10:46, 12-16-98
ENTERED

TD VOL= 2.5000 AT 10:47, 12-16-98
ENTERED

MF=-11.5 mV AT 10:51, 12-16-98

MF=-12.0 mV AT 10:51, 12-16-98

MF=-12.4 mV AT 10:52, 12-16-98

MF=-12.5 mV AT 10:52, 12-16-98

MF=-12.7 mV AT 10:53, 12-16-98
ENTERED

:NH3 SLOPE=-57.7 mV/DEC
AT 10:53, 12-16-98

:NH3 CONC= 2.40

3002 1 ml

DOUBLE KNOWN ADDITION SELECTED
AT 13:01, 12-16-98

SAMPLE VOL= 25.000 AT 13:02, 12-16-98
ENTERED

MF= 85.9 mV AT 13:09, 12-16-98

MF= 84.9 mV AT 13:10, 12-16-98

MF= 84.4 mV AT 13:10, 12-16-98

MF= 84.1 mV AT 13:10, 12-16-98
ENTERED

TD CONC= 1000 AT 13:11, 12-16-98
ENTERED

TD VOL= .25000 AT 13:11, 12-16-98
ENTERED

MF= 42.6 mV AT 13:16, 12-16-98

MF= 41.7 mV AT 13:16, 12-16-98

MF= 41.2 mV AT 13:17, 12-16-98

MF= 40.8 mV AT 13:17, 12-16-98

MF= 40.6 mV AT 13:18, 12-16-98

MF= 40.5 mV AT 13:18, 12-16-98
ENTERED

TD VOL= 2.5000 AT 13:18, 12-16-98
ENTERED

MF=-12.4 mV AT 13:23, 12-16-98

MF=-12.9 mV AT 13:23, 12-16-98

MF=-13.3 mV AT 13:24, 12-16-98

MF=-13.4 mV AT 13:24, 12-16-98
ENTERED

:NH3 SLOPE=-58.3 mV/DEC
AT 13:24, 12-16-98

:NH3 CONC= 2.15   3002 Dup 1ml

DOUBLE KNOWN ADDITION SELECTED
AT 13:28, 12-16-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 13:28, 12-16-98
ENTERED

MF= 144.8 mV AT 13:48, 12-16-98

MF= 145.8 mV AT 13:49, 12-16-98

MF= 146.6 mV AT 13:50, 12-16-98

MF= 147.0 mV AT 13:50, 12-16-98

MF= 147.3 mV AT 13:51, 12-16-98
ENTERED

TD CONC= 1000 AT 13:51, 12-16-98
ENTERED

TD VOL= .25000 AT 13:51, 12-16-98
ENTERED

MF= 42.9 mV AT 13:56, 12-16-98

MF= 42.6 mV AT 13:57, 12-16-98

MF= 42.3 mV AT 13:57, 12-16-98

MF= 42.3 mV AT 13:57, 12-16-98
ENTERED

TD VOL= 2.5000 AT 13:58, 12-16-98
ENTERED

MF=-14.2 mV AT 14:01, 12-16-98

MF=-14.8 mV AT 14:01, 12-16-98

MF=-15.1 mV AT 14:02, 12-16-98

MF=-15.4 mV AT 14:02, 12-16-98

MF=-15.5 mV AT 14:03, 12-16-98

MF=-15.6 mV AT 14:03, 12-16-98
ENTERED

:NH3 SLOPE=-58.2 mV/DEC
AT 14:03, 12-16-98

:NH3 CONC= .158   3003 1ml

DOUBLE KNOWN ADDITION SELECTED
AT 14:06, 12-16-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 14:06, 12-16-98
ENTERED

MF= 79.7 mV AT 14:12, 12-16-98

MF= 79.3 mV AT 14:12, 12-16-98

MF= 130.5 mV AT 14:28, 12-16-98

MF= 131.1 mV AT 14:28, 12-16-98

MF= 131.9 mV AT 14:29, 12-16-98

MF= 132.9 mV AT 14:30, 12-16-98

MF= 133.3 mV AT 14:30, 12-16-98

MF= 133.7 mV AT 14:30, 12-16-98

MF= 134.3 mV AT 14:31, 12-16-98

MF= 134.9 mV AT 14:32, 12-16-98

MF= 135.0 mV AT 14:32, 12-16-98

MF= 135.2 mV AT 14:32, 12-16-98
ENTERED

TD CONC= 1000 AT 14:32, 12-16-98
ENTERED

TD VOL= .25000 AT 14:33, 12-16-98
ENTERED

MF= 42.6 mV AT 14:38, 12-16-98

MF= 42.1 mV AT 14:38, 12-16-98

MF= 41.8 mV AT 14:39, 12-16-98

MF= 41.6 mV AT 14:39, 12-16-98
ENTERED

STD VOL= 2.5000 AT 14:39, 12-16-98
ENTERED

MF=-14.6 mV AT 14:44, 12-16-98

MF=-15.0 mV AT 14:44, 12-16-98

MF=-15.3 mV AT 14:45, 12-16-98

MF=-15.5 mV AT 14:45, 12-16-98
ENTERED

:NH3 SLOPE=-57.6 mV/DEC
AT 14:45, 12-16-98

:NH3 CONC= .241

3003 Dup 1ml

833

DOUBLE KNOWN ADDITION SELECTED
AT 14:50, 12-16-98

SAMPLE VOL= 25.000 AT 14:51, 12-16-98
ENTERED

EMF= 27.9 mV AT 14:52, 12-16-98

EMF= 27.5 mV AT 14:53, 12-16-98

EMF= 27.1 mV AT 14:53, 12-16-98

EMF= 26.9 mV AT 14:53, 12-16-98
ENTERED

STD CONC= 1000 AT 14:54, 12-16-98
ENTERED

STD VOL= .25000 AT 14:54, 12-16-98
ENTERED

EMF= 17.1 mV AT 14:57, 12-16-98

EMF= 16.8 mV AT 14:58, 12-16-98

EMF= 16.6 mV AT 14:58, 12-16-98

EMF= 16.5 mV AT 14:58, 12-16-98
ENTERED

STD VOL= 2.5000 AT 14:59, 12-16-98
ENTERED

EMF=-18.3 mV AT 15:03, 12-16-98

EMF=-18.7 mV AT 15:04, 12-16-98

EMF=-19.0 mV AT 15:04, 12-16-98

EMF=-19.3 mV AT 15:04, 12-16-98

EMF=-19.7 mV AT 15:05, 12-16-98

EMF=-19.7 mV AT 15:05, 12-16-98
ENTERED

:NH3 SLOPE=-60.9 mV/DEC
AT 15:05, 12-16-98

:NH3 CONC= 20.1

50N19-F 1ml

WORKBOOK PAGE: STD2

AMMONIA (NH3) : LA-631-001 (C-01) ¹⁻¹⁰ 3-25-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 18.700
STD	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
LIQUID		
Batch Number		
98005446		
Rerun#		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.67E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
12/30/98		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/16/98	QC FOUND (µg)	4.67E+02
Analysis Time		
03:30 PM		
Sample Point		
BY-105		

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 (C-01334)

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		12/31/98
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005446		
Rerun		
0		
Sample Prep		12/31/98
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.00E+00
Sample #		55
BLNK	Detection Limit (µg/mL)	5.00E+00
Instrument Code		55
NH301	Detection Limit = 0.200 µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
12/30/98		
Analysis Date		
12/16/98	0.75	
Analysis Time		
03:30 PM		12/31/98
Sample Point	NH3 Concentration (µg/mL)	< 5.00E+00
BY-105		55

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-~~6~~¹)₃₂₅₄₄

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 2.400
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005446		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 6.49E+02
Sample #		
S98T003002	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
12/30/98		
Analysis Date		
12/16/98		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	6.49E+02
BY-105		

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	<i>MF</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) ^{1/30/98}

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 2.150
DUP	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005446		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 5.80E+02
Sample #		
S98T003002	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
12/30/98		
Analysis Date		
12/16/98		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	5.80E+02
BY-105		

Analyst:	MF Date: 12/30/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM5

AMMONIA (NH3) : LA-631-001 (C-01)^{1, 2, 3, 4}

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.158
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005446		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.50E+01
Sample #		
S98T003003	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
12/30/98		
Analysis Date		
12/16/98		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	< 5.50E+01
BY-105		

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP6

AMMONIA (NH3) : LA-631-001 (C-0) 325-49

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.241
DUP	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005446		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 5.54E+01
Sample #		
S98T003003	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} \cdot (\text{FVOL}/\text{VSAM}) \cdot \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) \cdot (\text{FVOL} / \text{VSAM}) \cdot \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
12/30/98		
Analysis Date		
12/16/98		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	5.54E+01
BY-105		

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END7

AMMONIA (NH3) : LA-631-001 (C-~~631~~³³⁵⁻⁹⁹)

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 20.100
STD	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26842	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
LIQUID		
Batch Number		
98005446		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 5.02E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
12/30/98		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/16/98	QC FOUND (µg)	5.02E+02
Analysis Time		
03:30 PM		
Sample Point		
BY-105		

Analyst:	MF	Date: 12/30/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 26882

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 JMV

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	NH3-01	LIQUID	4.25E+02	4.39E+2	103.294 % Recovery	
2 BLNK	0	NH3-01	LIQUID	1	<1.00E+1		ug/mL
3 SAMPLE	S98T002956 0	NH3-01	LIQUID	N/A	<1.00E+01	10.000	ug/mL
4 DUP	S98T002956 0	NH3-01	LIQUID	<1.00E+1	<1.00E+1		RPD
5 SAMPLE	S98T002957 0	NH3-01	LIQUID	N/A	<1.00E+01	10.000	ug/mL
6 DUP	S98T002957 0	NH3-01	LIQUID	<1.00E+1	<1.00E+1		RPD
7 SPK	S98T002957 0	NH3-01	LIQUID	1.06E+02	1.06E+02	100.000 % Recovery	
8 STD	0	NH3-01	LIQUID	4.25E+02	4.62E+2	108.706 % Recovery	

Final page for worklist# 26882

Analyst Signature

¹¹⁻¹⁰⁻⁹⁸

Date

Analyst Signature Date

Reviewer Signature Date ^{11/17/98}

LABCORE Data Entry Template for Worklist# 26882

Analyst: MF Instrument: NH301 C-1 Book# 50119-D

Method: LA-631-001 Rev/Mod C1 10-25-99

Worklist Comment: BY-105 FOR NH3-01 JMV

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			NH3-01	LIQUID		
2 BLNK			NH3-01	LIQUID		
3 SAMPLE	S98T002956 0		NH3-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T002956 0		NH3-01	LIQUID		
5 SAMPLE	S98T002957 0		NH3-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: NH3-01					
6 DUP	S98T002957 0		NH3-01	LIQUID		
7 SPK	S98T002957 0		NH3-01	LIQUID		
8 STD			NH3-01	LIQUID		

Final page for worklist # 26882

Mary Fran 10-30-98
Analyst Signature Date 1545

Analyst Signature Date

Data Entry Comments:

DOUBLE KNOWN ADDITION SELECTED
AT 10:17, 10-30-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 10:17, 10-30-98
ENTERED

MF= 47.8 mV AT 10:21, 10-30-98

MF= 47.5 mV AT 10:21, 10-30-98

MF= 47.2 mV AT 10:21, 10-30-98

MF= 47.0 mV AT 10:21, 10-30-98

MF= 46.9 mV AT 10:22, 10-30-98

MF= 46.8 mV AT 10:22, 10-30-98

MF= 46.7 mV AT 10:23, 10-30-98

MF= 46.6 mV AT 10:23, 10-30-98

MF= 46.6 mV AT 10:24, 10-30-98

ENTERED

STD CONC= 1000 AT 10:24, 10-30-98
ENTERED

STD VOL= .25000 AT 10:24, 10-30-98
ENTERED

MF= 28.1 mV AT 10:26, 10-30-98

MF= 27.9 mV AT 10:27, 10-30-98

MF= 27.7 mV AT 10:27, 10-30-98

MF= 27.6 mV AT 10:28, 10-30-98

MF= 27.5 mV AT 10:28, 10-30-98

ENTERED

STD VOL= 2.5000 AT 10:29, 10-30-98
ENTERED

MF=-16.5 mV AT 10:31, 10-30-98

MF=-16.8 mV AT 10:31, 10-30-98

MF=-17.0 mV AT 10:32, 10-30-98

MF=-17.2 mV AT 10:32, 10-30-98

MF=-17.3 mV AT 10:33, 10-30-98

ENTERED

:NH3 SLOPE=-59.1 mV/DEC
AT 10:33, 10-30-98

:NH3 CONC= 8.88

.500ml 50N19-D

~~:NH3 SLOPE=-58.7 mV/DEC~~
~~AT 11:57, 10-30-98~~

:NH3 CONCEN= .210

DOUBLE KNOWN ADDITION SELECTED
AT 11:52, 10-30-98

SAMPLE VOL= 25.000 AT 11:52, 10-30-98
ENTERED

EMF= 156.5 mV AT 12:04, 10-30-98

EMF= 157.2 mV AT 12:04, 10-30-98

EMF= 157.6 mV AT 12:05, 10-30-98

EMF= 158.3 mV AT 12:05, 10-30-98

EMF= 158.7 mV AT 12:06, 10-30-98

EMF= 159.1 mV AT 12:06, 10-30-98
ENTERED

STD CONCEN= 1000 AT 12:06, 10-30-98
ENTERED

STD VOL= .25000 AT 12:07, 10-30-98
ENTERED

EMF= 41.2 mV AT 12:10, 10-30-98

EMF= 40.8 mV AT 12:11, 10-30-98

EMF= 40.8 mV AT 12:11, 10-30-98
ENTERED

STD VOL= 2.5000 AT 12:11, 10-30-98
ENTERED

EMF=-17.3 mV AT 12:14, 10-30-98

EMF=-17.8 mV AT 12:14, 10-30-98

EMF=-18.0 mV AT 12:15, 10-30-98
ENTERED

:NH3 SLOPE=-59.0 mV/DEC
AT 12:15, 10-30-98

:NH3 CONCEN= .0989

Bpk

DOUBLE KNOWN ADDITION SELECTED
AT 14:35, 10-30-98

SAMPLE VOL= 25.000 AT 14:35, 10-30-98
ENTERED

EMF= 148.4 mV AT 14:38, 10-30-98

EMF= 148.5 mV AT 14:38, 10-30-98

EMF= 146.0 mV AT 14:41, 10-30-98

EMF= 144.7 mV AT 14:42, 10-30-98

EMF= 144.0 mV AT 14:42, 10-30-98

EMF= 143.5 mV AT 14:42, 10-30-98

EMF= 142.2 mV AT 14:43, 10-30-98

EMF= 141.5 mV AT 14:44, 10-30-98

EMF= 141.2 mV AT 14:44, 10-30-98

ENTERED

STD CONC= 1000 AT 14:45, 10-30-98
ENTERED

STD VOL= .25000 AT 14:45, 10-30-98
ENTERED

EMF= 41.2 mV AT 14:47, 10-30-98

EMF= 41.0 mV AT 14:48, 10-30-98

EMF= 40.8 mV AT 14:48, 10-30-98

EMF= 40.8 mV AT 14:49, 10-30-98

ENTERED

STD VOL= 2.5000 AT 14:49, 10-30-98
ENTERED

EMF=-14.4 mV AT 14:50, 10-30-98

EMF=-18.3 mV AT 14:52, 10-30-98

EMF=-18.5 mV AT 14:53, 10-30-98

EMF=-18.5 mV AT 14:54, 10-30-98

EMF=-18.5 mV AT 14:54, 10-30-98

ENTERED

NH3 SLOPE=-59.8 mV/DEC
AT 14:54, 10-30-98

NH3 CONC= .211   2956 .500ml

DOUBLE KNOWN ADDITION SELECTED
AT 14:56, 10-30-98

AMPLE VOL= 25.000 AT 14:57, 10-30-98
ENTERED

EMF= 145.5 mV AT 15:08, 10-30-98

EMF= 146.0 mV AT 15:09, 10-30-98

EMF= 146.2 mV AT 15:09, 10-30-98
ENTERED

STD CONC= 1000 AT 15:09, 10-30-98
ENTERED

STD VOL= .25000 AT 15:09, 10-30-98
ENTERED

EMF= 42.3 mV AT 15:11, 10-30-98

EMF= 42.0 mV AT 15:12, 10-30-98

EMF= 41.8 mV AT 15:12, 10-30-98

EMF= 41.6 mV AT 15:13, 10-30-98
ENTERED

STD VOL= 2.5000 AT 15:13, 10-30-98
ENTERED

EMF=-16.6 mV AT 15:15, 10-30-98

EMF=-16.8 mV AT 15:15, 10-30-98

EMF=-17.1 mV AT 15:16, 10-30-98

EMF=-17.1 mV AT 15:16, 10-30-98
ENTERED

NH3 SLOPE=-59.1 mV/DEC
AT 15:16, 10-30-98

NH3 CONC= .171   .500ml 2956 Dup

DOUBLE KNOWN ADDITION SELECTED
AT 15:18, 10-30-98

SAMPLE VOL= 25.000 AT 15:18, 10-30-98
ENTERED

MF= 149.1 mV AT 15:30, 10-30-98

MF= 149.9 mV AT 15:30, 10-30-98

MF= 151.0 mV AT 15:31, 10-30-98

MF= 151.4 mV AT 15:31, 10-30-98
ENTERED

TD CONC= 1000 AT 15:32, 10-30-98
ENTERED

TD VOL= .25000 AT 15:32, 10-30-98
ENTERED

MF= 43.0 mV AT 15:34, 10-30-98

MF= 41.8 mV AT 15:35, 10-30-98

MF= 41.6 mV AT 15:35, 10-30-98

MF= 41.5 mV AT 15:36, 10-30-98
ENTERED

TD VOL= 2.5000 AT 15:36, 10-30-98
ENTERED

MF=-16.8 mV AT 15:38, 10-30-98

MF=-17.0 mV AT 15:39, 10-30-98

MF=-17.1 mV AT 15:39, 10-30-98
ENTERED

NH3 SLOPE=-58.9 mV/DEC
AT 15:39, 10-30-98

NH3 CONC= .137 .500mD 2957

DOUBLE KNOWN ADDITION SELECTED
AT 15:40, 10-30-98

SAMPLE VOL= 25.000 AT 15:40, 10-30-98
ENTERED

EMF= 151.0 mV AT 15:53, 10-30-98

EMF= 152.1 mV AT 15:53, 10-30-98

EMF= 154.0 mV AT 15:55, 10-30-98

EMF= 154.0 mV AT 15:55, 10-30-98
ENTERED

STD CONC= 1000 AT 15:55, 10-30-98
ENTERED

STD VOL= .25000 AT 15:55, 10-30-98
ENTERED

EMF= 44.7 mV AT 15:57, 10-30-98

EMF= 42.3 mV AT 15:58, 10-30-98

EMF= 42.2 mV AT 15:58, 10-30-98
ENTERED

STD VOL= 2.5000 AT 15:59, 10-30-98
ENTERED

EMF=-15.9 mV AT 16:01, 10-30-98

EMF=-16.3 mV AT 16:01, 10-30-98

EMF=-16.4 mV AT 16:01, 10-30-98
ENTERED

NH3 SLOPE=-58.9 mV/DEC
AT 16:01, 10-30-98

NH3 CONC= .126 *2957 Wap, 500ml*

DOUBLE KNOWN ADDITION SELECTED
AT 16:02, 10-30-98

SAMPLE VOL= 25.000 AT 16:04, 10-30-98
ENTERED

MF= 62.9 mV AT 16:07, 10-30-98

MF= 63.3 mV AT 16:07, 10-30-98

MF= 63.5 mV AT 16:07, 10-30-98
ENTERED

TD CONC= 1000 AT 16:08, 10-30-98
ENTERED

TD VOL= .25000 AT 16:08, 10-30-98
ENTERED

MF= 33.8 mV AT 16:10, 10-30-98

MF= 33.6 mV AT 16:10, 10-30-98

MF= 33.2 mV AT 16:11, 10-30-98
ENTERED

TD VOL= 2.5000 AT 16:11, 10-30-98
ENTERED

MF=-17.5 mV AT 16:13, 10-30-98

MF=-17.7 mV AT 16:13, 10-30-98
ENTERED

:NH3 SLOPE=-59.2 mV/DEC
AT 16:13, 10-30-98

:NH3 CONC= 4.38

.500ml 2957 spk .250ml 30N19-D

DOUBLE KNOWN ADDITION SELECTED

AT 16:28, 10-30-98

SAMPLE VOL= 25.000 AT 16:28, 10-30-98

ENTERED

MF= 45.7 mV AT 16:31, 10-30-98

MF= 46.0 mV AT 16:31, 10-30-98

MF= 46.1 mV AT 16:31, 10-30-98

ENTERED

TD CONC= 1000 AT 16:31, 10-30-98

ENTERED

TD VOL= .25000 AT 16:32, 10-30-98

ENTERED

MF= 27.3 mV AT 16:34, 10-30-98

MF= 27.2 mV AT 16:34, 10-30-98

ENTERED

TD VOL= 2.5000 AT 16:34, 10-30-98

ENTERED

MF=-17.8 mV AT 16:36, 10-30-98

MF=-18.1 mV AT 16:37, 10-30-98

MF=-18.2 mV AT 16:37, 10-30-98

ENTERED

..NH3 SLOPE=-60.6 mV/DEC

AT 16:37, 10-30-98

..NH3 CONC= 9.33

.500 ml 50N19.0

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-01) ¹ ₃₂₅₋₉₁

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 8.880
STD	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.39E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
10/30/98	QC FOUND (µg)	4.39E+02
Analysis Time		
03:45 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist: <i>MF</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: BLANK2

AMMONIA (NH3) : LA-631-001 (C-~~0~~¹)₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.099
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
BLK	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/30/98		
Analysis Time		
03:45 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:

MF Date: 11/10/98

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-~~0~~¹) 305-99

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.211
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002956	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/30/98		
Analysis Time		
03:45 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-D) ¹⁴⁰ ₃₂₅₋₉₉

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.171
DUP	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002956	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/30/98		
Analysis Time		
03:45 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:

MF Date: 11/10/98

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM5

AMMONIA (NH3) : LA-631-001 (C-0)^{1nd}₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.137
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002957	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/30/98		
Analysis Time		
03:45 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:

MF Date: 11/10/98

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP6

AMMONIA (NH3) : LA-631-001 (C-0) ^{مسجل} 3-25-99

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.126
DUP	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002957	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
10/30/98		
Analysis Time		
03:45 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END8

AMMONIA (NH3) : LA-631-001 (C-0) ¹³⁻⁰⁰⁻⁹⁹ _{LA-0}

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 9.330
STD	Blank Result from the Instrument (µg/mL)	BR 0.099
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26882	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005486		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.62E+02
Sample #		
	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
10/30/98	QC FOUND (µg)	4.62E+02
Analysis Time		
03:45 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 26883

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY 105 FOR NH3-01 JMV

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD	0		NH3-01	LIQUID	4.25E+02	4.17E+2	98.118 % Recovery	
2	BLNK	0		NH3-01	LIQUID	1	<1.00E+1	ug/mL	
3	SAMPLE	S98T002958	0	NH3-01	LIQUID	N/A	< 1.00E+01	10.000	ug/mL
4	DUP	S98T002958	0	NH3-01	LIQUID	<1.00E+1	<1.00E+1	RPD	
5	SPK	S98T002958	0	NH3-01	LIQUID	1.06E+02	1.01E+02	95.283 % Recovery	
6	SAMPLE	S98T002959	0	NH3-01	LIQUID	N/A	< 1.00E+01	10.000	ug/mL
7	DUP	S98T002959	0	NH3-01	LIQUID	<1.00E+1	<1.00E+1	RPD	
8	STD	0		NH3-01	LIQUID	4.25E+02	4.49E+2	105.647 % Recovery	

Final page for worklist# 26883

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

10/28/98 13:06
A-0004-1

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26883

Analyst: MF Instrument: NH301 3-25-99 Book# 50N19-D

Method: LA-631-001 Rev/Mod C-1 3-25-99

Worklist Comment: BY 105 FOR NH3-01 JMV

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				NH3-01	LIQUID		
2	BLNK				NH3-01	LIQUID		
3	SAMPLE	S98T002958 0			NH3-01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: NH3-01						
4	DUP	S98T002958 0			NH3-01	LIQUID		
5	SPK	S98T002958 0			NH3-01	LIQUID		
6	SAMPLE	S98T002959 0			NH3-01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: NH3-01						
7	DUP	S98T002959 0			NH3-01	LIQUID		
8	STD				NH3-01	LIQUID		

Final page for worklist # 26883

Mary Jane 11-2-98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

AMPLE VOL= 25.000 AT 09:59. 11-02-98
ENTERED

MF= 52.4 mV AT 10:03. 11-02-98

MF= 52.0 mV AT 10:03. 11-02-98

MF= 51.7 mV AT 10:04. 11-02-98

MF= 51.1 mV AT 10:05. 11-02-98

MF= 51.0 mV AT 10:05. 11-02-98

MF= 50.9 mV AT 10:06. 11-02-98

MF= 50.9 mV AT 10:06. 11-02-98
ENTERED

STD CONC= 1000 AT 10:06. 11-02-98
ENTERED

STD VOL= .25000 AT 10:06. 11-02-98
ENTERED

EMF= 31.5 mV AT 10:08. 11-02-98

EMF= 31.4 mV AT 10:09. 11-02-98

EMF= 31.0 mV AT 10:09. 11-02-98

EMF= 31.0 mV AT 10:09. 11-02-98
ENTERED

STD VOL= 2.5000 AT 10:09. 11-02-98
ENTERED

EMF=-6.3 mV AT 10:11. 11-02-98

EMF=-14.2 mV AT 10:13. 11-02-98

EMF=-14.5 mV AT 10:13. 11-02-98

EMF=-14.6 mV AT 10:14. 11-02-98

EMF=-14.6 mV AT 10:14. 11-02-98
ENTERED

E: NH3 SLOPE=-59.5 mV/DEC
AT 10:14. 11-02-98

E: NH3 CONC= 8.47

.500ml 5DN19-D

DOUBLE KNOWN ADDITION SELECTED

AT 10:17, 11-02-98

AMPLE VOL= 25.000 AT 10:17, 11-02-98

ENTERED

MF= 149.4 mV AT 10:27, 11-02-98

MF= 152.0 mV AT 10:28, 11-02-98

MF= 154.2 mV AT 10:30, 11-02-98

MF= 158.1 mV AT 10:33, 11-02-98

MF= 160.1 mV AT 10:35, 11-02-98

MF= 160.4 mV AT 10:35, 11-02-98

ENTERED

TD CONC= 1000 AT 10:35, 11-02-98

ENTERED

TD VOL= .25000 AT 10:35, 11-02-98

ENTERED

MF= 47.7 mV AT 10:38, 11-02-98

MF= 47.4 mV AT 10:38, 11-02-98

MF= 47.1 mV AT 10:39, 11-02-98

MF= 46.9 mV AT 10:39, 11-02-98

ENTERED

TD VOL= 2.5000 AT 10:40, 11-02-98

ENTERED

MF=-12.1 mV AT 10:42, 11-02-98

MF=-12.3 mV AT 10:42, 11-02-98

MF=-12.4 mV AT 10:43, 11-02-98

ENTERED

:NH3 SLOPE=-59.6 mV/DEC

AT 10:43, 11-02-98

:NH3 CONC= .125

Belk

DOUBLE KNOWN ADDITION SELECTED
AT 11:35. 11-02-98

SAMPLE VOL= 25.000 AT 11:36. 11-02-98
ENTERED

MF= 242.6 mV AT 11:36. 11-02-98

MF= 242.3 mV AT 11:36. 11-02-98

MF= 154.9 mV AT 12:01. 11-02-98

MF= 155.3 mV AT 12:01. 11-02-98

MF= 155.4 mV AT 12:01. 11-02-98
ENTERED

TD CONC= 1000 AT 12:02. 11-02-98
ENTERED

TD VOL= .25000 AT 12:02. 11-02-98
ENTERED

IMF= 47.2 mV AT 12:04. 11-02-98

IMF= 46.9 mV AT 12:05. 11-02-98

IMF= 46.9 mV AT 12:05. 11-02-98
ENTERED

STD VOL= 2.5000 AT 12:05. 11-02-98
ENTERED

EMF=-12.0 mV AT 12:07. 11-02-98

EMF=-12.2 mV AT 12:08. 11-02-98

EMF=-12.4 mV AT 12:09. 11-02-98

EMF=-12.4 mV AT 12:09. 11-02-98

EMF=-12.3 mV AT 12:09. 11-02-98
ENTERED

LNH3 SLOPE=-59.5 mV/DEC
AT 12:09. 11-02-98

LNH3 CONC= .151

.500 ml 5987002958

TUBLE KNOWN ADDITION SELECTED

AT 12:11, 11-02-98

AMPLE VOL= 25.000 AT 12:11, 11-02-98
ENTERED

MF= 155.3 mV AT 12:25, 11-02-98

MF= 156.4 mV AT 12:26, 11-02-98

MF= 157.6 mV AT 12:27, 11-02-98

MF= 157.6 mV AT 12:27, 11-02-98
ENTEREDTD CONC= 1000 AT 12:27, 11-02-98
ENTEREDTD VOL= .25000 AT 12:27, 11-02-98
ENTERED

MF= 48.0 mV AT 12:30, 11-02-98

MF= 47.6 mV AT 12:30, 11-02-98

MF= 47.5 mV AT 12:30, 11-02-98
ENTEREDTD VOL= 2.5000 AT 12:30, 11-02-98
ENTERED

MF=-11.7 mV AT 12:33, 11-02-98

MF=-12.0 mV AT 12:33, 11-02-98

MF=-12.0 mV AT 12:33, 11-02-98
ENTERED

NH3 SLOPE=-59.8 mV/DEC

AT 12:33, 11-02-98

NH3 CONC= .145

.500ml 2958 Dup

DOUBLE KNOWN ADDITION SELECTED

AT 14:05, 11-02-98

SAMPLE VOL= 25.000 AT 14:06, 11-02-98

ENTERED

MF= 68.5 mV AT 14:10, 11-02-98

MF= 68.3 mV AT 14:11, 11-02-98

MF= 68.1 mV AT 14:11, 11-02-98

MF= 68.0 mV AT 14:11, 11-02-98

ENTERED

TD CONC= 1000 AT 14:11, 11-02-98

ENTERED

TD VOL= .25000 AT 14:11, 11-02-98

ENTERED

MF= 38.0 mV AT 14:14, 11-02-98

MF= 37.8 mV AT 14:14, 11-02-98

MF= 37.5 mV AT 14:14, 11-02-98

MF= 37.4 mV AT 14:14, 11-02-98

ENTERED

TD VOL= 2.5000 AT 14:15, 11-02-98

ENTERED

MF=-12.7 mV AT 14:17, 11-02-98

MF=-13.0 mV AT 14:18, 11-02-98

ENTERED

:NH3 SLOPE=-58.3 mV/DEC

AT 14:18, 11-02-98

:NH3 CONC= 4.20

.500ml 2958 Spk, .250ml 50N19-D

DOUBLE KNOWN ADDITION SELECTED

AT 14:20. 11-02-98

AMPLE VOL= 25.000 AT 14:20. 11-02-98
ENTERED

MF= 139.4 mV AT 14:31. 11-02-98

MF= 140.2 mV AT 14:31. 11-02-98

MF= 141.3 mV AT 14:32. 11-02-98

MF= 141.5 mV AT 14:32. 11-02-98
ENTEREDTD CONC= 1000 AT 14:33. 11-02-98
ENTEREDTD VOL= .25000 AT 14:33. 11-02-98
ENTERED

MF= 47.6 mV AT 14:35. 11-02-98

MF= 47.3 mV AT 14:36. 11-02-98

MF= 47.2 mV AT 14:36. 11-02-98
ENTEREDTD VOL= 2.5000 AT 14:37. 11-02-98
ENTERED

MF=-11.3 mV AT 14:39. 11-02-98

MF=-11.5 mV AT 14:39. 11-02-98

MF=-11.5 mV AT 14:40. 11-02-98
ENTERED:NH3 SLOPE=-59.3 mV/DEC
AT 14:40. 11-02-98

:NH3 CONC= .261

. 500 ml 2959

DOUBLE KNOWN ADDITION SELECTED

AT 14:42, 11-02-98

AMPLE VOL= 25.000 AT 14:42, 11-02-98
ENTERED

MF= 149.8 mV AT 14:52, 11-02-98

MF= 151.6 mV AT 14:53, 11-02-98

MF= 151.7 mV AT 14:53, 11-02-98
ENTEREDSTD CONC= 1000 AT 14:53, 11-02-98
ENTEREDSTD VOL= .25000 AT 14:53, 11-02-98
ENTERED

EMF= 48.0 mV AT 14:56, 11-02-98

EMF= 47.6 mV AT 14:56, 11-02-98

EMF= 47.4 mV AT 14:57, 11-02-98
ENTEREDSTD VOL= 2.5000 AT 14:57, 11-02-98
ENTERED

EMF=-11.1 mV AT 14:59, 11-02-98

EMF=-11.3 mV AT 14:59, 11-02-98

EMF=-11.4 mV AT 15:00, 11-02-98
ENTERED1:NH3 SLOPE=-59.2 mV/DEC
AT 15:00, 11-02-98

1:NH3 CONC= .174

.500ml 2959 Dup

AMPLE KNOWN ADDITION SELECTED
AT 15:36, 11-02-98

AMPLE VOL= 25.000 AT 15:36, 11-02-98
ENTERED

MF= 48.2 mV AT 15:36, 11-02-98

MF= 48.9 mV AT 15:37, 11-02-98

MF= 49.5 mV AT 15:37, 11-02-98

MF= 49.7 mV AT 15:38, 11-02-98

MF= 50.0 mV AT 15:38, 11-02-98

MF= 50.3 mV AT 15:40, 11-02-98
ENTERED

TD CONC= 1000 AT 15:40, 11-02-98
ENTERED

TD VOL= .25000 AT 15:40, 11-02-98
ENTERED

MF= 31.5 mV AT 15:42, 11-02-98

MF= 31.1 mV AT 15:43, 11-02-98

MF= 31.1 mV AT 15:43, 11-02-98
ENTERED

TD VOL= 2.5000 AT 15:44, 11-02-98
ENTERED

MF=-14.0 mV AT 15:46, 11-02-98

MF=-14.4 mV AT 15:46, 11-02-98

MF=-14.5 mV AT 15:46, 11-02-98
ENTERED

:NH3 SLOPE=-60.5 mV/DEC
AT 15:46, 11-02-98

:NH3 CONC= 9.11 .500ml 50N19-D

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-6) ¹⁴⁰ 305-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 8.470
STD	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
liquid		
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.17E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
11/02/98	QC FOUND (µg)	4.17E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

AMMONIA (NH3) : LA-631-001 (C-0) 3-25-99

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.125
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
BLK	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/02/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH₃) : LA-631-001 (C-0) ^{1.40} 30541

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.151
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002958	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/02/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF	Date: 11/10/98
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-~~0~~¹)₃₋₅¹⁴⁰

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.145
DUP	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002958	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/02/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (C-~~0~~⁴⁰⁰) 3-25-99

LIQUIDS/SOLIDS

		SPK
Type	Instrument Data (µg/mL)	ID 4.200
SPK	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	Spike Book Number	SPK 50N19D
NH3-01	Spike Value (µg/mL)	SPK VAL 4.25E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK 0.250
LIQUID		
Batch Number		
98005487		
Rerun	Sample Instrument Data (µg/mL)	SAM ID 0.151
0	Sample Volume of Sample (mL)	SAM VSAM 0.500
Sample Prep	Sample Final Volume (mL)	SAM FVOL 25.0
N/A		
Sample #		
S98T002958		
Instrument Code		
NH301		
Prepared By		
JDS		
Chemist		
MJL	QC ACTUAL = SPK VAL * VOL SPK	
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	1.06E+02
11/02/98	QC FOUND (µg)	1.01E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

873

WORKBOOK PAGE: SAM6

AMMONIA (NH3) : LA-631-001 (C-0) ¹²⁰ 3-25-99

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.261
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002959	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/02/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: DUP7

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ¹⁴³ ³²⁵⁴⁹

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.174
DUP	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
S98T002959	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/02/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: ST_END8

AMMONIA (NH3) : LA-631-001 (C-~~0~~)^{1 µg} 3-25-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 9.110
STD	Blank Result from the Instrument (µg/mL)	BR 0.125
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26883	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005487		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.49E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
11/02/98	QC FOUND (µg)	4.49E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF	Date: 11/10/98
Signature of Chemist:	MJL	Date:
SAMPLE.WB1 REV 1.0	631001ML	

LABCORE Completed Worklist Report for Worklist# 26885

Analyst: krm Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	NH3-01	LIQUID	3.98E+02	4.53E+2	113.819 % Recovery	
2 BLNK	0	NH3-01	LIQUID	1	<5.00E+0		ug/mL
3 SAMPLE	S98T002812 0	NH3-01	LIQUID	N/A	6.33E+01	55.000	ug/mL
4 DUP	S98T002812 0	NH3-01	LIQUID	6.33E+1	2.20E+2	110.625 RPD	
5 SAMPLE	S98T002813 0	NH3-01	LIQUID	N/A	2.96E+02	55.000	ug/mL
6 DUP	S98T002813 0	NH3-01	LIQUID	2.96E+2	3.05E+2	2.995 RPD	
7 STD	0	NH3-01	LIQUID	3.98E+02	4.68E+2	117.588 % Recovery	

Final page for worklist# 26885

Analyst Signature _____ Date _____

Y. Mary Jones 1-11-99
Analyst Signature _____ Date _____

[Signature] 1/14/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 26885

Analyst: KRM Instrument: NH301 Book# 50N19F

Method: LA-631-001 Rev/Mod C-01 ¹⁻²⁵⁻⁹⁹

Worklist Comment: BY-105 FOR NH3-01 RTS

Prepare 1-10mL dilution. Use 1.5 mL sample size
12/16/98

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			NH3-01	LIQUID		
2 BLNK			NH3-01	LIQUID		
3 SAMPLE	S98T002812 0		NH3-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: NH3-01						
4 DUP	S98T002812 0		NH3-01	LIQUID		
5 SAMPLE	S98T002813 0		NH3-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: NH3-01						
6 DUP	S98T002813 0		NH3-01	LIQUID		
7 STD			NH3-01	LIQUID		

Final page for worklist # 26885

[Signature] 1-1-99
Analyst Signature Date

[Signature] 1-11-99
Analyst Signature Date

Data Entry Comments:

1-10 dilution / 1 ml SS used

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

SAMPLE VOL= 25.000 AT 00:26, 01-02-98
ENTERED

HNF-1663 REV. 0

EMF= 30.9 mV AT 00:29, 01-02-98

EMF= 31.3 mV AT 00:29, 01-02-98

EMF= 31.9 mV AT 00:30, 01-02-98

EMF= 32.1 mV AT 00:30, 01-02-98

EMF= 32.4 mV AT 00:31, 01-02-98

EMF= 32.8 mV AT 00:32, 01-02-98

EMF= 32.9 mV AT 00:32, 01-02-98
ENTERED

STD CONC= 1000 AT 00:32, 01-02-98
ENTERED

STD VOL= .25000 AT 00:32, 01-02-98
ENTERED

EMF= 22.0 mV AT 00:34, 01-02-98

EMF= 21.7 mV AT 00:34, 01-02-98

EMF= 21.7 mV AT 00:34, 01-02-98
ENTERED

STD VOL= 2.5000 AT 00:35, 01-02-98
ENTERED

EMF=-15.5 mV AT 00:37, 01-02-98

EMF=-15.8 mV AT 00:37, 01-02-98

EMF=-16.1 mV AT 00:38, 01-02-98

EMF=-16.1 mV AT 00:38, 01-02-98
ENTERED

1:NH3 SLOPE=-62.1 mV/DEC
AT 00:38, 01-02-98

1:NH3 CONC= 18.9 /

DOUBLE KNOWN ADDITION SELECTED
AT 00:38, 01-02-98

STD

SAMPLE VOL= 25.000 AT 23:33, 01-01-98
ENTERED

EMF= 85.6 mV AT 23:35, 01-01-98

EMF= 95.0 mV AT 23:39, 01-01-98

EMF= 95.3 mV AT 23:39, 01-01-98

EMF= 95.5 mV AT 23:40, 01-01-98

EMF= 95.9 mV AT 23:40, 01-01-98

EMF= 96.2 mV AT 23:41, 01-01-98

EMF= 96.4 mV AT 23:41, 01-01-98

EMF= 96.4 mV AT 23:42, 01-01-98
ENTERED

STD CONC= 1000 AT 23:42, 01-01-98
ENTERED

STD VOL= .25000 AT 23:42, 01-01-98
ENTERED

EMF= 42.7 mV AT 23:44, 01-01-98

EMF= 42.2 mV AT 23:46, 01-01-98

EMF= 42.3 mV AT 23:46, 01-01-98
ENTERED

STD VOL= 2.5000 AT 23:47, 01-01-98
ENTERED

EMF=-11.5 mV AT 23:48, 01-01-98

EMF=-12.1 mV AT 23:48, 01-01-98

EMF=-12.4 mV AT 23:49, 01-01-98

EMF=-12.5 mV AT 23:49, 01-01-98
ENTERED

1:NH3 SLOPE=-57.5 μ V/DEC
AT 23:49, 01-01-98

1:NH3 CONC= 1.28

DOUBLE KNOWN ADDITION SELECTED
AT 23:49, 01-01-98

2813 D4P

SAMPLE VOL= 25.000 AT 22:47, 01-01-98
ENTERED

HNF-1663 REV. 0

EMF= 98.1 mV AT 22:49, 01-01-98

EMF= 98.2 mV AT 22:50, 01-01-98

EMF= 98.7 mV AT 22:50, 01-01-98

EMF= 99.1 mV AT 22:51, 01-01-98

EMF= 99.1 mV AT 22:51, 01-01-98
ENTERED

STD CONC= 1000 AT 22:51, 01-01-98
ENTERED

STD VOL= .25000 AT 22:51, 01-01-98
ENTERED

EMF= 44.6 mV AT 22:53, 01-01-98

EMF= 44.4 mV AT 22:53, 01-01-98

EMF= 44.2 mV AT 22:54, 01-01-98
ENTERED

STD VOL= 2.5000 AT 22:54, 01-01-98
ENTERED

EMF=-10.2 mV AT 22:56, 01-01-98

EMF=-10.5 mV AT 22:56, 01-01-98

EMF=-10.7 mV AT 22:56, 01-01-98

EMF=-10.9 mV AT 22:57, 01-01-98

EMF=-11.1 mV AT 22:58, 01-01-98

EMF=-11.0 mV AT 22:58, 01-01-98
ENTERED

1:NH3 SLOPE=-57.9 mV/DEC
AT 22:58, 01-01-98

1:NH3 CONC= 1.25,

DOUBLE KNOWN ADDITION SELECTED
AT 22:58, 01-01-98

9985 2813

SAMPLE VOL= 25.000 AT 22:18, 01-01-98
ENTERED

EMF= 109.7 mV AT 22:21, 01-01-98

EMF= 110.0 mV AT 22:21, 01-01-98

EMF= 110.4 mV AT 22:22, 01-01-98

EMF= 110.8 mV AT 22:22, 01-01-98

EMF= 111.0 mV AT 22:23, 01-01-98

EMF= 111.2 mV AT 22:23, 01-01-98

EMF= 111.2 mV AT 22:24, 01-01-98
ENTERED

STD CONC= 1000 AT 22:24, 01-01-98
ENTERED

STD VOL= .25000 AT 22:24, 01-01-98
ENTERED

EMF= 49.9 mV AT 22:27, 01-01-98

EMF= 49.6 mV AT 22:27, 01-01-98

EMF= 49.4 mV AT 22:27, 01-01-98

EMF= 49.4 mV AT 22:28, 01-01-98
ENTERED

STD VOL= 2.5000 AT 22:28, 01-01-98
ENTERED

EMF=-6.7 mV AT 22:30, 01-01-98

EMF=-7.0 mV AT 22:31, 01-01-98

EMF=-7.3 mV AT 22:31, 01-01-98

EMF=-7.4 mV AT 22:31, 01-01-98
ENTERED

1:NH3 SLOPE=-58.9 mV/DEC
AT 22:31, 01-01-98

1:NH3 CONC= .971

DOUBLE KNOWN ADDITION SELECTED
AT 22:32, 01-01-98

2812 Dup

SAMPLE VOL= 25.000 AT 20:24, 01-01-98
ENTERED

EMF= 127.2 mV AT 20:28, 01-01-98

EMF= 127.8 mV AT 20:29, 01-01-98

EMF= 128.5 mV AT 20:30, 01-01-98

EMF= 128.7 mV AT 20:30, 01-01-98
ENTERED

STD CONC= 1000 AT 20:30, 01-01-98
ENTERED

STD VOL= .25000 AT 20:30, 01-01-98
ENTERED

EMF= 48.6 mV AT 20:31, 01-01-98

EMF= 48.4 mV AT 20:32, 01-01-98

EMF= 48.5 mV AT 20:33, 01-01-98

EMF= 48.5 mV AT 20:33, 01-01-98

EMF= 48.6 mV AT 20:34, 01-01-98

EMF= 48.8 mV AT 20:34, 01-01-98

EMF= 48.9 mV AT 20:35, 01-01-98
ENTERED

STD VOL= 2.5000 AT 20:35, 01-01-98
ENTERED

EMF=-6.5 mV AT 20:36, 01-01-98

EMF=-6.7 mV AT 20:36, 01-01-98

EMF=-6.8 mV AT 20:37, 01-01-98

EMF=-6.9 mV AT 20:37, 01-01-98
ENTERED

1:NH3 SLOPE=-56.7 mV/DEC
AT 20:37, 01-01-98

1:NH3 CONC= .402

DOUBLE KNOWN ADDITION SELECTED
AT 20:37, 01-01-98

598T 2812

SAMPLE VOL= 25.000 AT 19:56, 01-01-98
ENTERED

HNF-1663 REV. 0

EMF= 150.5 mV AT 20:02, 01-01-98

EMF= 151.2 mV AT 20:02, 01-01-98

EMF= 151.8 mV AT 20:03, 01-01-98

EMF= 151.9 mV AT 20:03, 01-01-98
ENTERED

BCK

STD CONC= 1000 AT 20:03, 01-01-98
ENTERED

STD VOL= .25000 AT 20:03, 01-01-98
ENTERED

EMF= 52.2 mV AT 20:04, 01-01-98

EMF= 51.9 mV AT 20:05, 01-01-98

EMF= 51.8 mV AT 20:05, 01-01-98

EMF= 51.8 mV AT 20:06, 01-01-98
ENTERED

STD VOL= 2.5000 AT 20:06, 01-01-98
ENTERED

EMF=-4.0 mV AT 20:07, 01-01-98

EMF=-4.3 mV AT 20:07, 01-01-98

EMF=-4.5 mV AT 20:08, 01-01-98

EMF=-4.5 mV AT 20:08, 01-01-98
ENTERED

1:NH3 SLOPE=-56.7 mV/DEC
AT 20:08, 01-01-98

1:NH3 CONC= .172

DOUBLE KNOWN ADDITION SELECTED
AT 20:08, 01-01-98

SAMPLE VOL= 25.000 AT 20:10, 01-01-98
ENTERED

HNF-1663 REV. 0

EMF= 37.5 mV AT 20:11, 01-01-98

EMF= 37.8 mV AT 20:11, 01-01-98

EMF= 38.0 mV AT 20:12, 01-01-98

EMF= 38.3 mV AT 20:13, 01-01-98

EMF= 38.4 mV AT 20:13, 01-01-98

EMF= 38.5 mV AT 20:14, 01-01-98

EMF= 38.6 mV AT 20:14, 01-01-98
ENTERED

STD

STD CONC= 1000 AT 20:14, 01-01-98
ENTERED

STD VOL= .25000 AT 20:15, 01-01-98
ENTERED

EMF= 27.6 mV AT 20:15, 01-01-98

EMF= 27.5 mV AT 20:16, 01-01-98

EMF= 27.6 mV AT 20:17, 01-01-98

EMF= 27.5 mV AT 20:17, 01-01-98
ENTERED

STD VOL= 2.5000 AT 20:17, 01-01-98
ENTERED

EMF=-8.9 mV AT 20:18, 01-01-98

EMF=-9.1 mV AT 20:18, 01-01-98

EMF=-9.2 mV AT 20:19, 01-01-98

EMF=-9.4 mV AT 20:20, 01-01-98

EMF=-9.3 mV AT 20:20, 01-01-98

EMF=-9.4 mV AT 20:20, 01-01-98
ENTERED

1:NH3 SLOPE=-59.9 mV/DEC
AT 20:21, 01-01-98

1:NH3 CONC= 18.3

DOUBLE KNOWN ADDITION SELECTED
AT 20:21, 01-01-98

HNF-1663
Rwo

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁰⁰ 32599		LIQUIDS/SOLIDS	STD
Type	Instrument Data (µg/mL)	ID	18.300
STD	Blank Result from the Instrument (µg/mL)	BR	0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	3.98E+02
Matrix			
LIQUID			
Batch Number			
98005489			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	4.53E+02
Sample #			
STD	Detection Limit (µg/mL)	5.00E+00	
Instrument Code			
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM)		
Prepared By			
MF	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)		
KRM			
Date Complete			
01/11/99			
Analysis Date	QC ACTUAL (µg)		3.98E+02
01/01/99	QC FOUND (µg)		4.53E+02
Analysis Time			
12:50 AM			
Sample Point			
BY-105			

Analyst:		KRM Date:	01/11/99
Signature of Chemist:	<i>NA</i>	MJL Date:	

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663
RWD

WORKBOOK PAGE: BLANK2

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁴⁰ 3-28-99		LIQUIDS/SOLIDS	BLNK
Type	Instrument Data (µg/mL)	ID	0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR	0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code			
NH3-01			
Matrix			
LIQUID	Dilution Factor	DF	1.000
Batch Number			
98005489			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	< 5.00E+00
Sample #			
BLNK	Detection Limit (µg/mL)	5.00E+00	
Instrument Code			
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM)		
Prepared By			
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF		
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
01/11/99			
Analysis Date			
01/01/99			
Analysis Time			
12:50 AM			
Sample Point	NH3 Concentration (µg/mL)	< 5.00E+00	
BY-105			

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

887

HWF-1663
Rwo

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁴⁰ 320899

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.402
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005489		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 6.33E+01
Sample #		
S98T002812	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
KRM		
Date Complete		
01/11/99		
Analysis Date		
01/01/99		
Analysis Time		
12:50 AM		
Sample Point	NH3 Concentration (µg/mL)	6.33E+01
BY-105		

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663
RWD

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-1) ¹ ~~1325~~ ⁹⁹ LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.971
DUP	Blank Result from the Instrument (µg/mL)	BR 0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005489		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.20E+02
Sample #		
S98T002812	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
KRM		
Date Complete		
01/11/99		
Analysis Date		
01/01/99		
Analysis Time		
12:50 AM		
Sample Point	NH3 Concentration (µg/mL)	2.20E+02
BY-105		

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663
RWO

WORKBOOK PAGE: SAM5

AMMONIA (NH3) : LA-631-001 (C-0) 2511

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 1.250
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005489		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.96E+02
Sample #		
S98T002813	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
KRM		
Date Complete		
01/11/99		
Analysis Date		
01/01/99		
Analysis Time		
12:50 AM		
Sample Point	NH3 Concentration (µg/mL)	2.96E+02
BY-105		

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663
R200

WORKBOOK PAGE: DUP6

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁻¹⁴⁻⁹⁹ 30541

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 1.280
DUP	Blank Result from the Instrument (µg/mL)	BR 0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005489		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.05E+02
Sample #		
S98T002813	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
KRM		
Date Complete		
01/11/99		
Analysis Date		
01/01/99		
Analysis Time		
12:50 AM		
Sample Point	NH3 Concentration (µg/mL)	3.05E+02
BY-105		

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663
Rev0

WORKBOOK PAGE: ST_END7

AMMONIA (NH3) : LA-631-001 (C-6) 3-25-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 18.900
STD	Blank Result from the Instrument (µg/mL)	BR 0.172
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26885	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
LIQUID		
Batch Number		
98005489		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.68E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH3-01	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
KRM		
Date Complete		
01/11/99		
Analysis Date	QC ACTUAL (µg)	3.98E+02
01/01/99	QC FOUND (µg)	4.68E+02
Analysis Time		
12:50 AM		
Sample Point		
BY-105		

Analyst:	KRM Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 26886

Analyst: rwk Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	STD	0		NH3-01	LIQUID	4.25E+02	4.19E+2	98.588 % Recovery	
2	BLNK	0		NH3-01	LIQUID	1	<5.00E+0	ug/mL	
3	SAMPLE	S98T002938	0	NH3-01	LIQUID	N/A	5.59E+01	5.000	ug/mL
4	DUP	S98T002938	0	NH3-01	LIQUID	5.59E+1	6.14E+1	9.378	RPD
5	SPK	S98T002938	0	NH3-01	LIQUID	2.13E+02	2.00E+02	93.897 % Recovery	
6	STD	0		NH3-01	LIQUID	4.25E+02	4.04E+2	95.059 % Recovery	

Final page for worklist# 26886

11/16/98
_____
Analyst Signature Date

_____
Analyst Signature Date

11/17/98
_____
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26886

Analyst: JK Instrument: NH301 Book# 50714-P

Method: LA-631-001 Rev/Mod C-01 12/3/99

Worklist Comment: BY-105 FOR NH3-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				NH3-01	LIQUID		
2	BLNK				NH3-01	LIQUID		
3	SAMPLE	S98T002938 0			NH3-01	LIQUID	98000490	BY-105 (2)
		Analytes Requested: NH3-01						
4	DUP	S98T002938 0			NH3-01	LIQUID		
5	SPK	S98T002938 0			NH3-01	LIQUID		
6	STD				NH3-01	LIQUID		

Final page for worklist # 26886

JK 11/14/98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 20:48. 11-03-98

SAMPLE VOL= 25.000 AT 20:48. 11-03-98
ENTERED

EMF= 36.3 mV AT 20:51. 11-03-98

EMF= 36.1 mV AT 20:51. 11-03-98

EMF= 35.8 mV AT 20:52. 11-03-98
ENTERED

STD CONC= 1000 AT 20:52. 11-03-98
ENTERED

STD VOL= .25000 AT 20:52. 11-03-98
ENTERED

EMF= 24.7 mV AT 20:53. 11-03-98

EMF= 24.4 mV AT 20:54. 11-03-98
ENTERED

STD VOL= 2.5000 AT 20:54. 11-03-98
ENTERED

EMF=-11.8 mV AT 20:56. 11-03-98

EMF=-12.0 mV AT 20:57. 11-03-98
ENTERED

std

1:NH3 SLOPE=-57.4 mV/DEC
AT 20:57. 11-03-98

1:NH3 CONC= 16.8

DOUBLE KNOWN ADDITION SELECTED
AT 20:57. 11-03-98

SAMPLE VOL= 25.000 AT 22:11. 11-03-98
ENTERED

EMF= 179.5 mV AT 22:11. 11-03-98

EMF= 179.2 mV AT 22:11. 11-03-98
ENTERED

STD CONC= 1000 AT 22:11. 11-03-98
ENTERED

STD VOL= .25000 AT 22:11. 11-03-98
ENTERED

EMF= 45.7 mV AT 22:15. 11-03-98

EMF= 45.5 mV AT 22:15. 11-03-98
ENTERED

STD VOL= 2.5000 AT 22:15. 11-03-98
ENTERED

BLK

EMF=-13.8 mV AT 22:19. 11-03-98

EMF=-13.8 mV AT 22:19. 11-03-98
ENTERED

1:NH3 SLOPE=-59.4 mV/DEC
AT 22:19. 11-03-98

1:NH3 CONC= .0559

DOUBLE KNOWN ADDITION SELECTED
AT 22:19. 11-03-98

SAMPLE VOL= 25.000 AT 22:41. 11-03-98
ENTERED

EMF= 84.3 mV AT 22:45. 11-03-98

EMF= 84.6 mV AT 22:46. 11-03-98

EMF= 84.9 mV AT 22:46. 11-03-98
ENTERED

STD CONC= 1000 AT 22:46. 11-03-98
ENTERED

STD VOL= .25000 AT 22:46. 11-03-98
ENTERED

EMF= 42.3 mV AT 22:49. 11-03-98

EMF= 42.2 mV AT 22:49. 11-03-98

EMF= 42.1 mV AT 22:49. 11-03-98
ENTERED

STD VOL= 2.5000 AT 22:50. 11-03-98
ENTERED

EMF=-12.1 mV AT 22:52. 11-03-98

EMF=-12.2 mV AT 22:53. 11-03-98
ENTERED

1:NH3 SLOPE=-59.0 mV/DEC
AT 22:53. 11-03-98

1:NH3 CONC= 2.29

597500 2938

1mL

DOUBLE KNOWN ADDITION SELECTED
AT 22:53. 11-03-98

SAMPLE VOL= 25.000 AT 23:21. 11-03-98
ENTERED

EMF= 87.2 mV AT 23:22. 11-03-98

EMF= 87.2 mV AT 23:22. 11-03-98

EMF= 87.2 mV AT 23:22. 11-03-98
ENTERED

STD CONC= 1000 AT 23:23. 11-03-98
ENTERED

STD VOL= .25000 AT 23:23. 11-03-98
ENTERED

EMF= 46.3 mV AT 23:25. 11-03-98

EMF= 46.2 mV AT 23:25. 11-03-98
ENTERED

STD VOL= 2.5000 AT 23:26. 11-03-98
ENTERED

EMF=-7.4 mV AT 23:28. 11-03-98

EMF=-7.7 mV AT 23:28. 11-03-98

EMF=-7.8 mV AT 23:28. 11-03-98
ENTERED

1:NH3 SLOPE=-59.2 mV/DEC
AT 23:29. 11-03-98

1:NH3 CONC= 2.51

598T002938 DUP

/ml

DOUBLE KNOWN ADDITION SELECTED
AT 23:30, 11-03-98

SAMPLE VOL= 25.000 AT 23:35, 11-03-98
ENTERED

EMF= 52.3 mV AT 23:36, 11-03-98

EMF= 52.2 mV AT 23:36, 11-03-98
ENTERED

STD CONC= 1000 AT 23:37, 11-03-98
ENTERED

STD VOL= .25000 AT 23:37, 11-03-98
ENTERED

EMF= 35.4 mV AT 23:39, 11-03-98

EMF= 35.2 mV AT 23:40, 11-03-98

EMF= 35.2 mV AT 23:40, 11-03-98
ENTERED

STD VOL= 2.5000 AT 23:40, 11-03-98
ENTERED

EMF=-7.3 mV AT 23:42, 11-03-98

EMF=-7.6 mV AT 23:42, 11-03-98
ENTERED

1:NH3 SLOPE=-58.5 mV/DEC
AT 23:42, 11-03-98

1:NH3 CONC= 10.3

5985002938 BPK
1ML
500ML BPK

DOUBLE KNOWN ADDITION SELECTED
AT 23:43. 11-03-98

SAMPLE VOL= 25.000 AT 23:47. 11-03-98
ENTERED

EMF= 44.2 mV AT 23:50. 11-03-98

EMF= 43.9 mV AT 23:50. 11-03-98

EMF= 43.8 mV AT 23:50. 11-03-98
ENTERED

STD CONC= 1000 AT 23:50. 11-03-98
ENTERED

STD VOL= .25000 AT 23:50. 11-03-98
ENTERED

EMF= 32.1 mV AT 23:52. 11-03-98

EMF= 31.9 mV AT 23:53. 11-03-98

EMF= 31.8 mV AT 23:53. 11-03-98
ENTERED

STD VOL= 2.5000 AT 23:53. 11-03-98
ENTERED

EMF=-5.4 mV AT 23:56. 11-03-98

EMF=-5.8 mV AT 23:56. 11-03-98

EMF=-5.9 mV AT 23:56. 11-03-98
ENTERED

1:NH3 SLOPE=-58.8 mV/DEC
AT 23:56, 11-03-98

1:NH3 CONC= 16.2

Ending Std

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁰⁰ 38544

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 16.800
STD	Blank Result from the Instrument (µg/mL)	BR 0.056
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
LIQUID		
Batch Number		
98005490		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.19E+02
Sample #		
STD	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
RWK		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
11/04/98	QC FOUND (µg)	4.19E+02
Analysis Time		
12:01 AM		
Sample Point		
BY105		

Analyst:	RWK Date: 11/10/98
Signature of Chemist: <i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: BLANK2

AMMONIA (NH₃) : LA-631-001 (C-0) ¹⁴⁰ 3-25-99

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.056
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.056
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005490		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.00E+00
Sample #		
BLK	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
RWK		
Date Complete		
11/10/98		
Analysis Date		
11/04/98		
Analysis Time		
12:01 AM		
Sample Point	NH3 Concentration (µg/mL)	< 5.00E+00
BY105		

Analyst:	RWK Date: 11/10/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

AMMONIA (NH3) : LA-631-001 (C-0) 3-25-99

LIQUIDS/SOLIDS

SAMPLE

Type	Instrument Data (µg/mL)	ID	2.290
SAMPLE	Blank Result from the Instrument (µg/mL)	BR	0.056
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	1.000
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code			
NH3-01			
Matrix			
LIQUID	Dilution Factor	DF	1.000
Batch Number			
98005490			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	5.59E+01
Sample #			
S98T002938	Detection Limit (µg/mL)	5.00E+00	
Instrument Code			
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF		
Prepared By			
JDS	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF		
Chemist			
MJL			
Analyst			
RWK			
Date Complete			
11/10/98			
Analysis Date			
11/04/98			
Analysis Time			
12:01 AM			
Sample Point	NH3 Concentration (µg/mL)		5.59E+01
BY105			

Analyst:		RWK	Date:	11/10/98
Signature of Chemist:	<i>NA</i>	MJL	Date:	

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-01) *32549*

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 2.510
DUP	Blank Result from the Instrument (µg/mL)	BR 0.056
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005490		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 6.14E+01
Sample #		
S98T002938	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
RWK		
Date Complete		
11/10/98		
Analysis Date		
11/04/98		
Analysis Time		
12:01 AM		
Sample Point	NH3 Concentration (µg/mL)	6.14E+01
BY105		

Analyst:		RWK Date: 11/10/98
Signature of Chemist:	<i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁻¹⁰ ₃₂₈₉

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	SPK
SPK	Blank Result from the Instrument (µg/mL)	BR	0.056
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	1.000
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	Spike Book Number	SPK	50N19D
NH3-01	Spike Value (µg/mL)	SPK VAL	4.25E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK	0.500
LIQUID			
Batch Number			
98005490			
Rerun	Sample Instrument Data (µg/mL)	SAM ID	2.29
0	Sample Volume of Sample (mL)	SAM VSAM	1.000
Sample Prep	Sample Final Volume (mL)	SAM FVOL	25.0
N/A			
Sample #			
S98T002938			
Instrument Code			
NH301			
Prepared By			
JDS			
Chemist			
MJL	QC ACTUAL = SPK VAL * VOL SPK		
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))		
RWK			
Date Complete			
11/10/98			
Analysis Date	QC ACTUAL (µg)		2.13E+02
11/04/98	QC FOUND (µg)		2.00E+02
Analysis Time			
12:01 AM			
Sample Point			
BY105			

Analyst:	RWK Date: 11/10/98
Signature of Chemist: <i>MA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END6

AMMONIA (NH3) : LA-631-001 (C-0) ¹₃₂₅₉

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	16.200
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.056
26886	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	1.000
Test Code	LCS Standard Book Number	LCS	25.0
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	50N19D
Matrix			4.25E+02
Batch Number			
98005490			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	4.04E+02
Sample #			
	Detection Limit (µg/mL)		5.00E+00
Instrument Code			
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$		
Prepared By			
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
RWK			
Date Complete			
11/10/98			
Analysis Date	QC ACTUAL (µg)		4.25E+02
11/04/98	QC FOUND (µg)		4.04E+02
Analysis Time			
12:01 AM			
Sample Point			
BY105			

Analyst:

RWK Date: 11/10/98

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Data Entry Template for Worklist# 26887

Analyst: RIM Instrument: NH301 Book# 50N19F

Method: LA-631-001 Rev/Mod C-01 ^{MAP} 3-25-99

Worklist Comment: BY-105 FOR NH3-01 RTS

Prepare 1-10mL dilution - Use 1.5mL 1.0mL RIM 12/31/98
sample size
Done 12/16/98

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	BLNK				NH3-01	LIQUID		
2	STD				NH3-01	LIQUID		
3	SAMPLE	S98T003004 0			NH3-01	LIQUID	98000510	BY-105 (2)
Analytes Requested: NH3-01								
4	DUP	S98T003004 0			NH3-01	LIQUID		
5	SPK	S98T003004 0			NH3-01	LIQUID		
6	SAMPLE	S98T003005 0			NH3-01	LIQUID	98000510	BY-105 (2)
Analytes Requested: NH3-01								
7	DUP	S98T003005 0			NH3-01	LIQUID		
* Trip					NH3-01	LIQUID		
8	STD				NH3-01	LIQUID		

Final page for worklist # 26887

RIM 12/31/98
Analyst Signature Date

Mary Frank 1-4-99
Analyst Signature Date

Data Entry Comments:

S98T003005 ran in triplicate

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 23:26, 12-30-98

STD 50N19F
1.0 mL

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 23:27, 12-30-98
ENTERED

RJ McCown

EMF= 40.6 mV AT 23:28, 12-30-98

EMF= 40.6 mV AT 23:28, 12-30-98
ENTERED

STD CONCN= 1000 AT 23:28, 12-30-98
ENTERED

STD VOL= .25000 AT 23:29, 12-30-98
ENTERED

EMF= 28.8 mV AT 23:29, 12-30-98

EMF= 28.8 mV AT 23:29, 12-30-98
ENTERED

STD VOL= 2.5000 AT 23:30, 12-30-98
ENTERED

EMF=-7.6 mV AT 23:31, 12-30-98

EMF=-7.6 mV AT 23:31, 12-30-98
ENTERED

1:NH3 SLOPE=-56.1 mV/DEC
AT 23:31, 12-30-98

1:NH3 CONCN= 15.6

Blank

DOUBLE KNOWN ADDITION SELECTED
AT 23:11, 12-30-98

SAMPLE VOL= 25.000 AT 23:12, 12-30-98
ENTERED

EMF= 201.3 mV AT 23:16, 12-30-98

EMF= 200.5 mV AT 23:17, 12-30-98

EMF= 200.5 mV AT 23:17, 12-30-98
ENTERED

STD CONCN= 1000 AT 23:17, 12-30-98
ENTERED

STD VOL= .25000 AT 23:17, 12-30-98
ENTERED

EMF= 58.0 mV AT 23:19, 12-30-98

EMF= 58.0 mV AT 23:19, 12-30-98
ENTERED

STD VOL= 2.5000 AT 23:20, 12-30-98
ENTERED

EMF=-1.1 mV AT 23:22, 12-30-98

EMF=-1.1 mV AT 23:22, 12-30-98
ENTERED

1:NH3 SLOPE=-59.2 mV/DEC
AT 23:23, 12-30-98

1:NH3 CONCN= .0388

5987003004 1-0 mL - 10 mL
used 1-0 mL
DOUBLE KNOWN ADDITION SELECTED
AT 01:00, 12-31-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 01:01, 12-31-98
ENTERED

EMF= 148.2 mV AT 01:03, 12-31-98

EMF= 148.2 mV AT 01:03, 12-31-98
ENTERED

STD CONC= 1000 AT 01:04, 12-31-98
ENTERED

STD VOL= .25000 AT 01:04, 12-31-98
ENTERED

EMF= 52.5 mV AT 01:06, 12-31-98

EMF= 52.4 mV AT 01:06, 12-31-98
ENTERED

STD VOL= 2.5000 AT 01:06, 12-31-98
ENTERED

EMF=-4.9 mV AT 01:08, 12-31-98

EMF=-5.1 mV AT 01:08, 12-31-98
ENTERED

1:NH3 SLOPE=-58.0 mV/DEC
AT 01:08, 12-31-98

1:NH3 CONC= .226

598T003004 Dup 1.0mL - 10mL
hold 1.0mL

HNF-1663 REV. 0

DOUBLE KNOWN ADDITION SELECTED

AT 01:20, 12-31-98

SAMPLE VOL= 25.000 AT 01:20, 12-31-98

ENTERED

EMF= 139.3 mV AT 01:23, 12-31-98

EMF= 139.5 mV AT 01:24, 12-31-98

ENTERED

STD CONC= 1000 AT 01:24, 12-31-98

ENTERED

STD VOL= .25000 AT 01:24, 12-31-98

ENTERED

EMF= 47.7 mV AT 01:25, 12-31-98

EMF= 47.6 mV AT 01:25, 12-31-98

ENTERED

STD VOL= 2.5000 AT 01:26, 12-31-98

ENTERED

EMF=-7.9 mV AT 01:27, 12-31-98

EMF=-8.0 mV AT 01:27, 12-31-98

ENTERED

1:NH3 SLOPE=-56.1 mV/DEC

AT 01:27, 12-31-98

1:NH3 CONC= .233

5987003004 Spk 1.0 mL - 10 mL
used 1.0 mL
DOUBLE KNOWN ADDITION SELECTED
AT 01:34, 12-31-98 +. 500 ML 50N19F

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 01:34, 12-31-98
ENTERED

EMF= 52.6 mV AT 01:35, 12-31-98

EMF= 52.5 mV AT 01:35, 12-31-98
ENTERED

STD CONC= 1000 AT 01:36, 12-31-98
ENTERED

STD VOL= .25000 AT 01:36, 12-31-98
ENTERED

EMF= 34.5 mV AT 01:37, 12-31-98

EMF= 34.5 mV AT 01:37, 12-31-98
ENTERED

STD VOL= 2.5000 AT 01:38, 12-31-98
ENTERED

EMF=-9.3 mV AT 01:39, 12-31-98

EMF=-9.4 mV AT 01:39, 12-31-98
ENTERED

1:NH3 SLOPE=-59.0 mV/DEC
AT 01:39, 12-31-98

1:NH3 CONC= 9.63

STD 50N19F 1.0mL

DOUBLE KNOWN ADDITION SELECTED
AT 03:47, 12-31-98

Resume Conn

HNF-1663 REV.0

AMPLE VOL= 25.000 AT 03:47, 12-31-98
ENTERED

EMF= 36.3 mV AT 03:49, 12-31-98

EMF= 36.2 mV AT 03:49, 12-31-98
ENTERED

STD CONC= 1000 AT 03:49, 12-31-98
ENTERED

STD VOL= .25000 AT 03:49, 12-31-98
ENTERED

EMF= 25.4 mV AT 03:51, 12-31-98

EMF= 25.3 mV AT 03:51, 12-31-98
ENTERED

STD VOL= 2.5000 AT 03:51, 12-31-98
ENTERED

EMF=-11.4 mV AT 03:53, 12-31-98

EMF=-11.5 mV AT 03:53, 12-31-98
ENTERED

1:NH3 SLOPE=-60.5 mV/DEC
AT 03:53, 12-31-98

1:NH3 CONC= 18.9

AMMONIA (NH3) : LA-631-001 (C-01) ¹ ₃₋₂₀₋₉₉ LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 15.600
STD	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
Batch Number		
98005491		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.89E+02
Sample #		
STD	Detection Limit (µg/mL)	0.00E+00
Instrument Code		
NH3-01	Detection Limit = $0.000\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
PJM		
Date Complete		
01/04/99		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/31/99	QC FOUND (µg)	3.89E+02
Analysis Time		
04:30 AM		
Sample Point		
BY105		

Analyst:	PJM Date: 01/04/99
Signature of Chemist: <i>N/A</i>	MJL Date:

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 (C-01)¹₃₂₅₉₉

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005491		
ReRun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 0.00E+00
Sample #		
BLNK	Detection Limit (µg/mL)	0.00E+00
Instrument Code		
NH3-01	Detection Limit = 0.000µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
PJM		
Date Complete		
01/04/99		
Analysis Date		
12/31/99		
Analysis Time		
04:30 AM		
Sample Point	NH3 Concentration (µg/mL)	< 0.00E+00
BY105		

Analyst:	PJM Date: 01/04/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0) ¹₃₂₅₄₉

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.226
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Main		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005491		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 5.15E+01
Sample #		
S98T003004	Detection Limit (µg/mL)	0.00E+00
Instrument Code		
NH3-01	Detection Limit = $0.000\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
PJM		
Date Complete		
01/04/99		
Analysis Date		
12/31/99		
Analysis Time		
04:30 AM		
Sample Point	NH3 Concentration (µg/mL)	5.15E+01
BY105		

Analyst:	PJM Date: 01/04/99
Signature of Chemist: <i>N/A</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) ^{14D} 325-99

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.233
DUP	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98005491		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 5.34E+01
Sample #		
S98T003004	Detection Limit (µg/mL)	0.00E+00
Instrument Code		
NH3-01	Detection Limit = $0.000\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
PJM		
Date Complete		
01/04/99		
Analysis Date		
12/31/99		
Analysis Time		
04:30 AM		
Sample Point	NH3 Concentration (µg/mL)	5.34E+01
BY105		

Analyst:	PJM Date: 01/04/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ₃₋₂₅₋₉₉ LIQUIDS/SOLIDS

		SPK
Type	Instrument Data (µg/mL)	ID 9.630
SPK	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	Spike Book Number	SPK 50N19F
NH3-01	Spike Value (µg/mL)	SPK VAL 3.98E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK 0.500
LIQUID		
Batch Number		
98005491		
Rerun	Sample Instrument Data (µg/mL)	SAM ID 0.226
0	Sample Volume of Sample (mL)	SAM VSAM 1.000
Sample Prep	Sample Final Volume (mL)	SAM FVOL 25.0
N/A		
Sample #		
S98T003004		
Instrument Code		
NH3-01		
Prepared By		
MF		
Chemist		
MJL	QC ACTUAL = SPK VAL * VOL SPK	
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))	
PJM		
Date Complete		
01/04/99		
Analysis Date	QC ACTUAL (µg)	1.99E+02
12/31/99	QC FOUND (µg)	2.35E+02
Analysis Time		
04:30 AM		
Sample Point		
BY105		

Analyst:	PJM Date: 01/04/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END9

AMMONIA (NH3) : LA-631-001 (C-0)¹_{3/24/99}

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 18.900
STD	Blank Result from the Instrument (µg/mL)	BR 0.039
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
26887	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
Batch Number		
98005491		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.72E+02
Sample #		
STD	Detection Limit (µg/mL)	0.00E+00
Instrument Code		
NH3-01	Detection Limit = 0.000µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
PJM		
Date Completed		
01/04/99		
Analysis Date	QC ACTUAL (µg)	3.98E+02
12/31/99	QC FOUND (µg)	4.72E+02
Analysis Time		
04:30 AM		
Sample Point		
BY105		

Analyst:		PJM Date: 01/04/99
Signature of Chemist:	<i>N/A</i>	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML	

LABCORE Completed Worklist Report for Worklist# 26890

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 FOR NH3-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	NH3-01	LIQUID	4.25E+02	3.89E+2	91.529 % Recovery	
2 BLNK	0	NH3-01	LIQUID	1	<1.00E+1		ug/mL
3 SAMPLE	S98T002808 0	NH3-01	LIQUID	N/A	1.17E+02	10.000	ug/mL
4 DUP	S98T002808 0	NH3-01	LIQUID	1.17E+2	1.05E+2	10.811 RPD	
5 SAMPLE	S98T002809 0	NH3-01	LIQUID	N/A	1.91E+02	10.000	ug/mL
6 DUP	S98T002809 0	NH3-01	LIQUID	1.91E+2	1.88E+2	1.583 RPD	
7 STD	0	NH3-01	LIQUID	4.25E+02	4.11E+2	96.706 % Recovery	

Final page for worklist# 26890

Analyst Signature Date 11/10/98

_____
Analyst Signature Date

Reviewer/Signature Date 11/17/98

LABCORE Data Entry Template for Worklist# 26890

Analyst: MF Instrument: NH301 Book# 50N19-D

Method: LA-631-001 Rev/Mod C-1

Worklist Comment: BY-105 FOR NH3-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				NH3-01	LIQUID		
2	BLNK				NH3-01	LIQUID		
3	SAMPLE	S98T002808 0			NH3-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: NH3-01						
4	DUP	S98T002808 0			NH3-01	LIQUID		
5	SAMPLE	S98T002809 0			NH3-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: NH3-01						
6	DUP	S98T002809 0			NH3-01	LIQUID		
7	STD				NH3-01	LIQUID		

Final page for worklist # 26890

Mary Tracy 11-5-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED

AT 10:06, 11-05-98

AMPLE VOL= 25.000 AT 10:06, 11-05-98
ENTERED

MF= 54.0 mV AT 10:09, 11-05-98

MF= 53.7 mV AT 10:10, 11-05-98

MF= 53.5 mV AT 10:10, 11-05-98

MF= 53.2 mV AT 10:11, 11-05-98

MF= 52.9 mV AT 10:11, 11-05-98

MF= 52.9 mV AT 10:11, 11-05-98
ENTEREDSTD CONC= 1000 AT 10:12, 11-05-98
ENTEREDSTD VOL= .25000 AT 10:12, 11-05-98
ENTERED

MF= 33.6 mV AT 10:14, 11-05-98

MF= 33.2 mV AT 10:15, 11-05-98

MF= 33.0 mV AT 10:15, 11-05-98

MF= 32.8 mV AT 10:16, 11-05-98

MF= 32.9 mV AT 10:17, 11-05-98
ENTEREDSTD VOL= 2.5000 AT 10:17, 11-05-98
ENTERED

MF=-10.8 mV AT 10:20, 11-05-98

MF=-11.0 mV AT 10:20, 11-05-98

MF=-11.2 mV AT 10:21, 11-05-98
ENTEREDNH3 SLOPE=-56.6 mV/DEC
AT 10:21, 11-05-98

NH3 CONC= 7.82 , 500ml 50N19-D

DOUBLE KNOWN ADDITION SELECTED

AT 11:27, 11-05-98

SAMPLE VOL= 25.000 AT 11:28, 11-05-98
ENTERED

EMF= 183.4 mV AT 11:51, 11-05-98

EMF= 189.1 mV AT 11:55, 11-05-98

EMF= 189.2 mV AT 11:55, 11-05-98
ENTEREDSTD CONC= 1000 AT 11:55, 11-05-98
ENTEREDSTD VOL= .25000 AT 11:55, 11-05-98
ENTERED

EMF= 45.4 mV AT 11:59, 11-05-98

EMF= 45.1 mV AT 11:59, 11-05-98

EMF= 44.9 mV AT 11:59, 11-05-98

EMF= 44.7 mV AT 12:00, 11-05-98

EMF= 44.7 mV AT 12:00, 11-05-98
ENTEREDSTD VOL= 2.5000 AT 12:00, 11-05-98
ENTERED

EMF=-14.3 mV AT 12:03, 11-05-98

EMF=-14.5 mV AT 12:03, 11-05-98

EMF=-14.6 mV AT 12:03, 11-05-98
ENTERED1: NH3 SLOPE=-59.4 mV/DEC
AT 12:03, 11-05-98

1: NH3 CONC= .0366

BLK

DOUBLE KNOWN ADDITION SELECTED

AT 12:04, 11-05-98

AMPLE VOL= 25.000 AT 12:06, 11-05-98
ENTERED

MF= 81.0 mV AT 12:12, 11-05-98

MF= 81.6 mV AT 12:13, 11-05-98

MF= 82.1 mV AT 12:13, 11-05-98

MF= 82.4 mV AT 12:14, 11-05-98

MF= 82.5 mV AT 12:14, 11-05-98

ENTERED

TD CONC= 1000 AT 12:14, 11-05-98

ENTERED

TD VOL= .25000 AT 12:14, 11-05-98

ENTERED

MF= 40.6 mV AT 12:18, 11-05-98

MF= 40.5 mV AT 12:18, 11-05-98

MF= 40.3 mV AT 12:18, 11-05-98

ENTERED

TD VOL= 2.5000 AT 12:19, 11-05-98

ENTERED

MF=-13.4 mV AT 12:21, 11-05-98

MF=-13.8 mV AT 12:22, 11-05-98

MF=-14.0 mV AT 12:22, 11-05-98

MF=-14.1 mV AT 12:22, 11-05-98

MF=-14.1 mV AT 12:23, 11-05-98

ENTERED

..NH3 SLOPE=-59.3 mV/DEC

AT 12:23, 11-05-98

..NH3 CONC= 2.38

398T002808 .500 ml

DOUBLE KNOWN ADDITION SELECTED
AT 14:16, 11-05-98

SAMPLE VOL= 25.000 AT 14:16, 11-05-98
ENTERED

MF= 82.4 mV AT 14:20, 11-05-98

MF= 81.9 mV AT 14:21, 11-05-98

MF= 81.8 mV AT 14:21, 11-05-98

MF= 81.7 mV AT 14:21, 11-05-98
ENTERED

TD CONC= 1000 AT 14:22, 11-05-98
ENTERED

TD VOL= .25000 AT 14:22, 11-05-98
ENTERED

MF= 37.7 mV AT 14:25, 11-05-98

MF= 37.4 mV AT 14:25, 11-05-98

MF= 37.2 mV AT 14:26, 11-05-98

MF= 36.9 mV AT 14:26, 11-05-98

MF= 36.9 mV AT 14:26, 11-05-98
ENTERED

STD VOL= 2.5000 AT 14:26, 11-05-98
ENTERED

MF=-17.7 mV AT 14:30, 11-05-98

MF=-18.1 mV AT 14:30, 11-05-98

MF=-18.4 mV AT 14:31, 11-05-98

MF=-18.3 mV AT 14:31, 11-05-98
ENTERED

NH3 SLOPE=-59.7 mV/DEC
AT 14:31, 11-05-98

NH3 CONC= 2.13 . 500 ml 2808 Dep

TABLE KNOWN ADDITION SELECTED
T 14:39, 11-05-98

MPLE VOL= 25.000 AT 14:39, 11-05-98
ENTERED

MF= 67.5 mV AT 14:45, 11-05-98

MF= 68.2 mV AT 14:46, 11-05-98

MF= 68.7 mV AT 14:46, 11-05-98

MF= 69.0 mV AT 14:46, 11-05-98

MF= 69.4 mV AT 14:47, 11-05-98

MF= 69.7 mV AT 14:47, 11-05-98

MF= 69.8 mV AT 14:47, 11-05-98

ENTERED

TD CONC= 1000 AT 14:48, 11-05-98
ENTERED

TD VOL= .25000 AT 14:48, 11-05-98
ENTERED

MF= 36.6 mV AT 14:51, 11-05-98

MF= 36.4 mV AT 14:51, 11-05-98

MF= 36.3 mV AT 14:52, 11-05-98
ENTERED

TD VOL= 2.5000 AT 14:52, 11-05-98
ENTERED

MF=-16.2 mV AT 14:55, 11-05-98

MF=-16.5 mV AT 14:56, 11-05-98

MF=-16.8 mV AT 14:57, 11-05-98
ENTERED

NH3 SLOPE=-60.8 mV/DEC
AT 14:57, 11-05-98

NH3 CONC= 3.86 *598T002809 .500 ml*

DOUBLE KNOWN ADDITION SELECTED

AT 14:59, 11-05-98

SAMPLE VOL= 25.000 AT 14:59, 11-05-98

ENTERED

MF= 69.3 mV AT 15:04, 11-05-98

MF= 70.1 mV AT 15:05, 11-05-98

MF= 70.5 mV AT 15:05, 11-05-98

MF= 70.8 mV AT 15:06, 11-05-98

MF= 71.0 mV AT 15:06, 11-05-98

MF= 71.2 mV AT 15:07, 11-05-98

MF= 71.3 mV AT 15:07, 11-05-98

MF= 71.6 mV AT 15:07, 11-05-98

MF= 71.5 mV AT 15:07, 11-05-98

ENTERED

TD CONC= 1000 AT 15:08, 11-05-98

ENTERED

TD VOL= .25000 AT 15:08, 11-05-98

ENTERED

MF= 37.9 mV AT 15:10, 11-05-98

MF= 37.7 mV AT 15:10, 11-05-98

MF= 37.5 mV AT 15:11, 11-05-98

MF= 37.5 mV AT 15:11, 11-05-98

ENTERED

TD VOL= 2.5000 AT 15:12, 11-05-98

ENTERED

MF=-15.6 mV AT 15:14, 11-05-98

MF=-15.8 mV AT 15:14, 11-05-98

MF=-15.9 mV AT 15:15, 11-05-98

MF=-16.0 mV AT 15:15, 11-05-98

ENTERED

:NH3 SLOPE=-61.1 mV/DEC

AT 15:15, 11-05-98

:NH3 CONC= 3.79

.500 ml 5987202809 disp

DOUBLE KNOWN ADDITION SELECTED

AT 15:28, 11-05-98

SAMPLE VOL= 25.000 AT 15:28, 11-05-98
ENTERED

MF= 51.6 mV AT 15:31, 11-05-98

MF= 51.2 mV AT 15:31, 11-05-98

MF= 51.0 mV AT 15:31, 11-05-98

MF= 50.8 mV AT 15:32, 11-05-98

MF= 50.6 mV AT 15:32, 11-05-98

MF= 50.6 mV AT 15:32, 11-05-98
ENTEREDED CONCEN= 1000 AT 15:32, 11-05-98
ENTEREDED VOL= .25000 AT 15:32, 11-05-98
ENTERED

MF= 30.9 mV AT 15:35, 11-05-98

MF= 30.7 mV AT 15:36, 11-05-98

MF= 30.6 mV AT 15:36, 11-05-98
ENTEREDED VOL= 2.5000 AT 15:37, 11-05-98
ENTERED

MF=-14.3 mV AT 15:40, 11-05-98

MF=-14.4 mV AT 15:40, 11-05-98

MF=-14.7 mV AT 15:41, 11-05-98

MF=-14.7 mV AT 15:41, 11-05-98
ENTEREDNH3 SLOPE=-58.8 mV/DEC
AT 15:41, 11-05-98

NH3 CONCEN= 8.26

.500ml 50N19D

WORKBOOK PAGE: STD1

AMMONIA (NH3) : LA-631-001 (C-0) ¹ 3-25-99 _{Lab} LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 7.820
STD	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.89E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
11/05/98	QC FOUND (µg)	3.89E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: BLANK2

AMMONIA (NH3) : LA-631-001 (C-0) ¹AD 3-25-99

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.037
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
BLK	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/05/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY105		

Analyst:	MF	Date: 11/10/98
Signature of Chemist:	<i>MF</i>	MJL Date:
SAMPLE.WB1 REV 1.0		
631001ML		

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0) 3-25-98

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 2.380
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.17E+02
Sample #		
S98T002808	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/05/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.17E+02
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) 14030599

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 2.130
DUP	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.05E+02
Sample #		
S98T002808	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/05/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.05E+02
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM5

AMMONIA (NH3) : LA-631-001 (C-0) ³⁻²⁵⁻⁹⁹ _{END}

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 3.860
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.91E+02
Sample #		
S98T002809	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JDS	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/05/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.91E+02
BY105		

Analyst:	MF Date: 11/10/98
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP6

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ^{MD} ₃₀₀₉₉

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 3.790
DUP	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.88E+02
Sample #		
S98T002809	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
11/10/98		
Analysis Date		
11/05/98		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.88E+02
BY105		

Analyst:	MF	Date: 11/10/98
Signature of Chemist:	<i>N/A</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END7

AMMONIA (NH₃) : LA-631-001 (C-0) ^{1 MD} ₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 8.260
STD	Blank Result from the Instrument (µg/mL)	BR 0.037
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
26890	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19D
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.25E+02
Matrix		
Batch Number		
98005494		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.11E+02
Sample #	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JDS	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
11/10/98		
Analysis Date	QC ACTUAL (µg)	4.25E+02
11/05/98	QC FOUND (µg)	4.11E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:		MF Date: 11/10/98
Signature of Chemist:		MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 27640

Analyst: rwk Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY-105 NH3-01 RLC

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	BLNK	0		NH3-01	LIQUID	1	<5.00E+0		ug/mL
2	STD	0		NH3-01	LIQUID	3.98E+02	4.04E+2	101.508 %	Recovery
3	SAMPLE	S98T003383	0	NH3-01	LIQUID	N/A	1.03E+03	55.000	ug/mL
4	DUP	S98T003383	0	NH3-01	LIQUID	1.03E+3	9.22E+1	167.136	RPD
5	SPK	S98T003383	0	NH3-01	LIQUID	1.99E+02	2.03E+02	102.010 %	Recovery
6	STD	0		NH3-01	LIQUID	3.98E+02	4.06E+2	102.010 %	Recovery

Final page for worklist# 27640

Analyst Signature _____ Date _____

Yancy Jones 1-11-99
Analyst Signature Date

[Signature] 1/14/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27640

Analyst: DLK Instrument: NH301 Book# 50114-FMethod: LA-631-001 Rev/Mod 2nd LSC CDC-1

Worklist Comment: BY-105 NH3-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			NH3-01	LIQUID		
2 STD			NH3-01	LIQUID		
3 SAMPLE	S98T003383 0		NH3-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T003383 0		NH3-01	LIQUID		
5 SPK	S98T003383 0		NH3-01	LIQUID		
6 STD			NH3-01	LIQUID		

Final page for worklist # 27640

Analyst Signature DLK Date 1/6/99Analyst Signature Mary Tracy Date _____

Data Entry Comments:

1-10 dilution / 1 ml SS used

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 14:11, 01-06-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 14:11, 01-06-98
ENTERED

EMF= 44.0 mV AT 14:18, 01-06-98

EMF= 42.3 mV AT 14:19, 01-06-98

EMF= 42.1 mV AT 14:19, 01-06-98
ENTERED

STD CONC= 1000 AT 14:20, 01-06-98
ENTERED

STD VOL= .25000 AT 14:20, 01-06-98
ENTERED

EMF=-64.9 mV AT 14:23, 01-06-98

EMF=-64.9 mV AT 14:23, 01-06-98
ENTERED

STD VOL= 2.5000 AT 14:24, 01-06-98
ENTERED

EMF=-123.3 mV AT 14:27, 01-06-98

EMF=-123.5 mV AT 14:27, 01-06-98
ENTERED

1:NH3 SLOPE=-58.9 mV/DEC
AT 14:27, 01-06-98

1:NH3 CONC= .154

7366

HNF-1663 REV. 0

DOUBLE KNOWN ADDITION SELECTED
AT 14:45, 01-06-98

SAMPLE VOL= 25.000 AT 14:45, 01-06-98
ENTERED

EMF=-74.3 mV AT 14:48, 01-06-98

EMF=-74.5 mV AT 14:48, 01-06-98

EMF=-74.5 mV AT 14:48, 01-06-98
ENTERED

STD CONC= 1000 AT 14:49, 01-06-98
ENTERED

STD VOL= .25000 AT 14:49, 01-06-98
ENTERED

EMF=-86.2 mV AT 14:51, 01-06-98

EMF=-86.4 mV AT 14:51, 01-06-98

EMF=-86.4 mV AT 14:51, 01-06-98
ENTERED

STD VOL= 2.5000 AT 14:52, 01-06-98
ENTERED

EMF=-123.8 mV AT 14:55, 01-06-98

EMF=-123.9 mV AT 14:55, 01-06-98
ENTERED

1:NH3 SLOPE=-58.6 mV/DEC
AT 14:55, 01-06-98

1:NH3 CONC= 15.3

DOUBLE KNOWN ADDITION SELECTED
AT 15:02, 01-06-98

SAMPLE VOL= 25.000 AT 15:02, 01-06-98
ENTERED

HNF-1663 REV. 0

EMF=-31.3 mV AT 15:03, 01-06-98

EMF=-31.2 mV AT 15:03, 01-06-98
ENTERED

STD CONC= 1000 AT 15:04, 01-06-98
ENTERED

STD VOL= .25000 AT 15:04, 01-06-98
ENTERED

EMF=-65.3 mV AT 15:07, 01-06-98

EMF=-65.4 mV AT 15:07, 01-06-98
ENTERED

STD VOL= 2.5000 AT 15:07, 01-06-98
ENTERED

EMF=-119.7 mV AT 15:11, 01-06-98

EMF=-119.8 mV AT 15:11, 01-06-98
ENTERED

1:NH3 SLOPE=-62.3 mV/DEC
AT 15:11, 01-06-98

1:NH3 CONC= 3.89

5487003383

DOUBLE KNOWN ADDITION SELECTED

AT 15:17, 01-06-98

SAMPLE VOL= 25.000 AT 15:17, 01-06-98
ENTERED

EMF=-28.7 mV AT 15:18, 01-06-98

EMF=-28.9 mV AT 15:19, 01-06-98

HNF-1663 REV. 0

EMF=-28.9 mV AT 15:19, 01-06-98
ENTERED

STD CONC= 1000 AT 15:19, 01-06-98
ENTERED

STD VOL= .25000 AT 15:19, 01-06-98
ENTERED

EMF=-62.9 mV AT 15:22, 01-06-98

EMF=-63.2 mV AT 15:23, 01-06-98

EMF=-63.2 mV AT 15:23, 01-06-98
ENTERED

STD VOL= 2.5000 AT 15:23, 01-06-98
ENTERED

3383 DUP.

EMF=-117.2 mV AT 15:27, 01-06-98

EMF=-117.5 mV AT 15:27, 01-06-98

EMF=-117.5 mV AT 15:27, 01-06-98
ENTERED

1:NH3 SLOPE=-62.1 mV/DEC
AT 15:27, 01-06-98

2:NH3 CONC= 3.84

DOUBLE KNOWN ADDITION SELECTED
AT 15:33, 01-06-98

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 15:33, 01-06-98
ENTERED

EMF=-55.9 mV AT 15:37, 01-06-98

EMF=-56.1 mV AT 15:37, 01-06-98

EMF=-56.8 mV AT 15:37, 01-06-98
ENTERED

STD CONC= 1000 AT 15:37, 01-06-98
ENTERED

STD VOL= .25000 AT 15:37, 01-06-98
ENTERED

EMF=-72.3 mV AT 15:40, 01-06-98

EMF=-72.5 mV AT 15:41, 01-06-98

EMF=-72.8 mV AT 15:41, 01-06-98

EMF=-72.8 mV AT 15:41, 01-06-98
ENTERED

STD VOL= 2.5000 AT 15:41, 01-06-98
ENTERED

EMF=-116.1 mV AT 15:46, 01-06-98

EMF=-116.3 mV AT 15:46, 01-06-98

EMF=-116.3 mV AT 15:46, 01-06-98
ENTERED

1:NH3 SLOPE=-61.9 mV/DEC
AT 15:46, 01-06-98

1:NH3 CONCN= 12.0

DOUBLE KNOWN ADDITION SELECTED
AT 15:47, 01-06-98

SAMPLE VOL= 25.000 AT 15:53, 01-06-98
ENTERED

EMF=-45.7 mV AT 15:53, 01-06-98

EMF=-66.6 mV AT 15:57, 01-06-98

EMF=-66.8 mV AT 15:57, 01-06-98
ENTERED

STD CONCN= 1000 AT 15:57, 01-06-98
ENTERED

STD VOL= .25000 AT 15:57, 01-06-98
ENTERED

EMF=-78.2 mV AT 16:00, 01-06-98

EMF=-79.1 mV AT 16:01, 01-06-98

EMF=-79.2 mV AT 16:01, 01-06-98
ENTERED

STD VOL= 2.5000 AT 16:01, 01-06-98
ENTERED

EMF=-118.0 mV AT 16:05, 01-06-98

EMF=-118.3 mV AT 16:05, 01-06-98

EMF=-118.3 mV AT 16:05, 01-06-98
ENTERED

1:NH3 SLOPE=-61.1 mV/DEC
AT 16:06, 01-06-98

1:NH3 CONCN= 16.4

HNF-1663 REV. 0

3383 SPK

Ending std

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ^{LD} 3-25-99

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98006263		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 5.00E+00
Sample #		
BLNK	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
RWK		
Date Complete		
01/11/99		
Analysis Date		
01/06/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 5.00E+00
BY-105		

Analyst:		RWK Date: 01/11/99
Signature of Chemist:	<i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: STD2

AMMONIA (NH3) : LA-631-001 (C-0) ¹ ~~2764~~ ²⁷⁶⁴

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL) ID	16.300
STD	Blank Result from the Instrument (µg/mL) BR	0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL) VSAM	1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL) FVOL	25.0
Test Code	LCS Standard Book Number LCS	50N19F
NH3-01	LCS Standard Concentration (µg/mL) STD VAL	3.98E+02
Matrix		
Batch Number		
98006263		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL) NH3 CONC	4.04E+02
Sample #		
STD	Detection Limit (µg/mL) 5.00E+00	
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
RWK		
Date Complete		
01/11/99		
Analysis Date	QC ACTUAL (µg)	3.98E+02
01/06/99	QC FOUND (µg)	4.04E+02
Analysis Time		
03:00 PM		
Sample Point		
BY-105		

Analyst:	RWK Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0) ³⁻²⁵⁻⁹⁹ ₁₄₃

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 3.890
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98006263		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.03E+03
Sample #		
S98T003383	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
RWK		
Date Complete		
01/11/99		
Analysis Date		
01/06/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.03E+03
BY-105		

Analyst:	RWK Date: 01/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) ¹⁴⁰ _{325.11}

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 3.840
DUP	Blank Result from the Instrument (µg/mL)	BR 0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 11.000
Batch Number		
98006263		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.01E+03
Sample #		
S98T003383	Detection Limit (µg/mL)	5.50E+01
Instrument Code		
NH301	Detection Limit = $0.200 \mu\text{g} * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID} - \text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
RWK		
Date Complete		
01/11/99		
Analysis Date		
01/06/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.01E+03
BY-105		

Analyst:		RWK Date: 01/11/99
Signature of Chemist:	<i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (C-~~0~~)⁴⁰ 3-25-99

LIQUIDS/SOLIDS

		SPK
Type	Instrument Data (µg/mL)	ID 12.000
SPK	Blank Result from the Instrument (µg/mL)	BR 0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	Spike Book Number	SPK 50N19F
NH3-01	Spike Value (µg/mL)	SPK VAL 3.98E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK 0.500
LIQUID		
Batch Number		
98006263		
Rerun	Sample Instrument Data (µg/mL)	SAM ID 3.89
0	Sample Volume of Sample (mL)	SAM VSAM 1.000
Sample Prep	Sample Final Volume (mL)	SAM FVOL 25.0
N/A		
Sample #		
S98T003383		
Instrument Code		
NH301		
Prepared By		
MF		
Chemist		
MJL	QC ACTUAL = SPK VAL * VOL SPK	
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))	
RWK		
Date Complete		
01/11/99		
Analysis Date	QC ACTUAL (µg)	1.99E+02
01/06/99	QC FOUND (µg)	2.03E+02
Analysis Time		
03:00 PM		
Sample Point		
BY-105		

Analyst:		RWK Date: 01/11/99
Signature of Chemist:		MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END6

AMMONIA (NH3) : LA-631-001 (C-0) ¹₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 16.400
STD	Blank Result from the Instrument (µg/mL)	BR 0.154
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 1.000
27640	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19F
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 3.98E+02
Matrix		
Batch Number		
98006263		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.06E+02
Sample #	Detection Limit (µg/mL)	5.00E+00
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
RWK		
Date Complete		
01/11/99		
Analysis Date	QC ACTUAL (µg)	3.98E+02
01/06/99	QC FOUND (µg)	4.06E+02
Analysis Time		
03:00 PM		
Sample Point		
BY-105		

Analyst:		RWK Date: 01/11/99
Signature of Chemist:	<i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 27835

Analyst: mf Instrument: NH302 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY105 FOR NH3-01 USE 1-10 DILUTION WITH 1ML S S MF jmv

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	NH3-01	LIQUID	1	<1.00E+1		ug/mL
2 STD	0	NH3-01	LIQUID	4.33E+02	4.24E+2	97.921 % Recovery	
3 SAMPLE	S98T003005 0	NH3-01	LIQUID	N/A	1.19E+02	10.000 ug/mL	
4 DUP	S98T003005 0	NH3-01	LIQUID	1.19E+2	1.14E+2	4.292 RPD	
5 STD	0	NH3-01	LIQUID	4.33E+02	3.42E+02	78.984 % Recovery	

Final page for worklist# 27835

Analyst Signature _____ Date _____

3-11-99
Analyst Signature _____ Date _____

3/15/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27835

Analyst: _____ Instrument: NH301 _____ Book# _____

Method: LA-631-001 Rev/Mod D-0

Worklist Comment: BY105 FOR NH3-01 USE 1-10 DILUTION WITH 1ML S S MF

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			NH3-01	LIQUID		
2 STD			NH3-01	LIQUID		
~~3~~ ¹⁰⁻⁹⁹ SAMPLE	~~S98T002010~~ 0		NH3-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: NH3-01					
~~4~~ ¹⁰⁻⁹⁹ DUP	~~S98T002810~~ 0		NH3-01	LIQUID		
5 SAMPLE	S98T003005 0		NH3-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: NH3-01					
6 DUP	S98T003005 0		NH3-01	LIQUID		
7 STD			NH3-01	LIQUID		

Final page for worklist # 27835

Analyst Signature _____ Date _____

Mary Tracy 3-10-99
Analyst Signature Date

Data Entry Comments:

Ran S98T003005 and Dup

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 08:34, 03-10-99

SAMPLE VOL= 25.000 AT 08:35, 03-10-99
ENTERED

MF= 64.8 mV AT 08:36, 03-10-99

MF= 63.8 mV AT 08:36, 03-10-99

MF= 82.3 mV AT 08:42, 03-10-99

MF= 121.4 mV AT 08:58, 03-10-99

MF= 122.8 mV AT 08:59, 03-10-99

MF= 123.5 mV AT 08:59, 03-10-99

MF= 125.0 mV AT 09:01, 03-10-99

MF= 125.1 mV AT 09:01, 03-10-99
ENTERED

STD CONC= 998 AT 09:01, 03-10-99
ENTERED

STD VOL= .25000 AT 09:01, 03-10-99
ENTERED

MF=-46.6 mV AT 09:04, 03-10-99

MF=-46.9 mV AT 09:04, 03-10-99
ENTERED

STD VOL= 2.5000 AT 09:05, 03-10-99
ENTERED

MF=-105.8 mV AT 09:07, 03-10-99

MF=-106.0 mV AT 09:07, 03-10-99

MF=-106.0 mV AT 09:07, 03-10-99
ENTERED

NH3 SLOPE=-59.1 mV/DEC
AT 09:07, 03-10-99

NH3 CONC= .0122   BLK

DOUBLE KNOWN ADDITION SELECTED
AT 09:24, 03-10-99

SAMPLE VOL= 25.000 AT 09:24, 03-10-99
ENTERED

MF=-14.0 mV AT 09:27, 03-10-99

MF=-14.3 mV AT 09:27, 03-10-99

MF=-14.6 mV AT 09:28, 03-10-99

MF=-14.7 mV AT 09:28, 03-10-99
ENTERED

STD CONC= 998 AT 09:28, 03-10-99
ENTERED

STD VOL= .25000 AT 09:28, 03-10-99
ENTERED

EMF=-34.9 mV AT 09:31, 03-10-99

EMF=-35.3 mV AT 09:31, 03-10-99

EMF=-35.5 mV AT 09:31, 03-10-99
ENTERED

STD VOL= 2.5000 AT 09:32, 03-10-99
ENTERED

EMF=-81.0 mV AT 09:34, 03-10-99

EMF=-81.5 mV AT 09:34, 03-10-99

EMF=-81.7 mV AT 09:35, 03-10-99

EMF=-81.9 mV AT 09:35, 03-10-99

EMF=-82.1 mV AT 09:35, 03-10-99

EMF=-82.3 mV AT 09:35, 03-10-99

EMF=-82.5 mV AT 09:36, 03-10-99

EMF=-82.7 mV AT 09:36, 03-10-99

EMF=-83.0 mV AT 09:37, 03-10-99

EMF=-83.1 mV AT 09:37, 03-10-99

EMF=-83.3 mV AT 09:37, 03-10-99
ENTERED

1:NH3 SLOPE=-62.5 mV/DEC
AT 09:37, 03-10-99

1:NH3 CONC= 8.50

.500 ml 50 N 19-H

DOUBLE KNOWN ADDITION SELECTED
AT 13:24, 03-10-99

SAMPLE VOL= 25.000 AT 13:24, 03-10-99
ENTERED

MF= 57.6 mV AT 13:30, 03-10-99

MF= 57.0 mV AT 13:30, 03-10-99

MF= 57.2 mV AT 13:31, 03-10-99

MF= 56.6 mV AT 13:31, 03-10-99
ENTERED

STD CONC= 998 AT 13:31, 03-10-99
ENTERED

STD VOL= .25000 AT 13:31, 03-10-99
ENTERED

EMF= 15.7 mV AT 13:34, 03-10-99

EMF= 14.9 mV AT 13:35, 03-10-99

EMF= 13.7 mV AT 13:35, 03-10-99

EMF= 13.4 mV AT 13:36, 03-10-99

EMF= 12.7 mV AT 13:36, 03-10-99

EMF= 12.3 mV AT 13:36, 03-10-99

EMF= 12.0 mV AT 13:37, 03-10-99

EMF= 11.6 mV AT 13:37, 03-10-99

EMF= 11.4 mV AT 13:37, 03-10-99
ENTERED

STD VOL= 2.5000 AT 13:38, 03-10-99
ENTERED

EMF=-45.2 mV AT 13:42, 03-10-99

EMF=-46.3 mV AT 13:43, 03-10-99

EMF=-46.9 mV AT 13:43, 03-10-99

EMF=-47.1 mV AT 13:44, 03-10-99
ENTERED

1:NH3 SLOPE=-63.9 mV/DEC
AT 13:44, 03-10-99

1:NH3 CONC= 2.40 , 500 ml 3005

DOUBLE KNOWN ADDITION SELECTED
AT 13:59, 03-10-99

AMPLE VOL= 25.000 AT 13:59, 03-10-99
ENTERED

MF= 48.9 mV AT 14:03, 03-10-99

IMF= 48.0 mV AT 14:03, 03-10-99

TMF= 47.2 mV AT 14:04, 03-10-99

IMF= 46.8 mV AT 14:04, 03-10-99

IMF= 46.6 mV AT 14:05, 03-10-99

IMF= 46.4 mV AT 14:05, 03-10-99

IMF= 46.3 mV AT 14:05, 03-10-99

ENTERED

STD CONC= 998 AT 14:06, 03-10-99
ENTERED

STD VOL= .25000 AT 14:06, 03-10-99
ENTERED

EMF= 4.0 mV AT 14:10, 03-10-99

EMF= 3.7 mV AT 14:10, 03-10-99

EMF= 2.9 mV AT 14:11, 03-10-99

EMF= 2.8 mV AT 14:11, 03-10-99

EMF= 2.5 mV AT 14:11, 03-10-99

EMF= 2.4 mV AT 14:12, 03-10-99

ENTERED

STD VOL= 2.5000 AT 14:12, 03-10-99
ENTERED

EMF=-52.3 mV AT 14:16, 03-10-99

EMF=-52.6 mV AT 14:16, 03-10-99

EMF=-52.9 mV AT 14:16, 03-10-99

EMF=-53.2 mV AT 14:17, 03-10-99

EMF=-53.3 mV AT 14:17, 03-10-99

ENTERED

1: NH3 SLOPE=-60.6 mV/DEC
AT 14:17, 03-10-99

2: NH3 CONC= 2.29 .500 *u* 3005

DOUBLE KNOWN ADDITION SELECTED

AT 14:56, 03-10-99

SAMPLE VOL= 25.000 AT 14:56, 03-10-99
ENTERED

EMF= 29.2 mV AT 15:01, 03-10-99

MF= 28.6 mV AT 15:01, 03-10-99

MF= 28.2 mV AT 15:02, 03-10-99

MF= 27.9 mV AT 15:02, 03-10-99

MF= 27.3 mV AT 15:03, 03-10-99

MF= 27.2 mV AT 15:03, 03-10-99

ENTERED

STD CONC= 998 AT 15:03, 03-10-99

ENTERED

STD VOL= .25000 AT 15:03, 03-10-99

ENTERED

EMF= 4.1 mV AT 15:06, 03-10-99

EMF= 3.8 mV AT 15:06, 03-10-99

EMF= 3.5 mV AT 15:07, 03-10-99

EMF= 3.3 mV AT 15:07, 03-10-99

ENTERED

STD VOL= 2.5000 AT 15:07, 03-10-99

ENTERED

EMF=-43.2 mV AT 15:10, 03-10-99

EMF=-43.7 mV AT 15:11, 03-10-99

EMF=-44.1 mV AT 15:11, 03-10-99

EMF=-44.7 mV AT 15:11, 03-10-99

EMF=-45.6 mV AT 15:12, 03-10-99

EMF=-46.1 mV AT 15:13, 03-10-99

EMF=-46.2 mV AT 15:13, 03-10-99

ENTERED

NH3 SLOPE=-61.9 mV/DEC

AT 15:13, 03-10-99

NH3 CONC= 6.85

. 500 ml 50N19-H

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 (C-0) ^D ^{LAD} ^{#2544}

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL) ID	0.000
BLNK	Blank Result from the Instrument (µg/mL) BR	0.012
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL) VSAM	0.500
27835	Final Vol of Dilution (mL) or Vol of Sample Direct (mL) FVOL	25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor DF	1.000
Batch Number		
99000081		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL) NH3 CONC	< 1.00E+01
Sample #		
BLNK	Detection Limit (µg/mL) 1.00E+01	
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/11/99		
Analysis Date		
03/10/99		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY 105		

Analyst:	MF Date: 03/11/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: STD2

AMMONIA (NH3) : LA-631-001 (C-0) ^{no} 325 91

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	0.012
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.500
27835	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	4.33E+02
Matrix			
Batch Number			
99000081			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	4.24E+02
Sample #			
	Detection Limit (µg/mL)		1.00E+01
Instrument Code			
NH302	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$		
Prepared By			
JMV	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
MF			
Date Complete			
03/11/99			
Analysis Date	QC ACTUAL (µg)		4.33E+02
03/10/99	QC FOUND (µg)		4.24E+02
Analysis Time			
03:30 PM			
Sample Point			
BY 105			

Analyst:	MF	Date:	03/11/99
Signature of Chemist:	<i>NA</i>	MJL	Date:
SAMPLE.WB1 REV 1.0		631001ML	

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (C-0) ^D ₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 2.400
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.012
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
27835	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000081		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.19E+02
Sample #		
S98T003005	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JMV	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/11/99		
Analysis Date		
03/10/99		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	1.19E+02
BY 105		

Analyst:	MF Date: 03/11/99
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (0-0) ^{D 3-25-99 LND}

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 2.290
DUP	Blank Result from the Instrument (µg/mL)	BR 0.012
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
27835	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000081		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.14E+02
Sample #		
S98T003005	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
JMV	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/11/99		
Analysis Date		
03/10/99		
Analysis Time		
03:30 PM		
Sample Point	NH3 Concentration (µg/mL)	1.14E+02
BY 105		

Analyst:	MF Date: 03/11/99
Signature of Chemist: <i>NA</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: ST_END5

AMMONIA (NH3) : LA-631-001 (C-0) ^{LAD 3-25-99}

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 6.850
STD	Blank Result from the Instrument (µg/mL)	BR 0.012
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
27835	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.33E+02
Matrix		
Batch Number		
99000081		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.42E+02
Sample #	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
03/11/99		
Analysis Date	QC ACTUAL (µg)	4.33E+02
03/10/99	QC FOUND (µg)	3.42E+02
Analysis Time		
03:30 PM		
Sample Point		
BY 105		

Analyst:

MF Date: 03/11/99

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

LABCORE Completed Worklist Report for Worklist# 27849

Analyst: mf Instrument: NH302 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY 105 NH3-01 USE 1-10 DILUTION WITH 1ML SS MF

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLANK	0	NH3-01	LIQUID	1	<2.00E+1		ug/mL
2 STD	0	NH3-01	LIQUID	4.33E+02	4.29E+2	99.076 % Recovery	
3 SAMPLE	S98T002811 0	NH3-01	LIQUID	N/A	2.77E+02	20.000 ug/mL	
4 DUP	S98T002811 0	NH3-01	LIQUID	2.77E+2	3.19E+2	14.094 RPD	
5 SPK	S98T002811 0	NH3-01	LIQUID	1.08E+02	6.97E+01	64.537 % Recovery	
6 SAMPLE	S98T002810 0	NH3-01	LIQUID	N/A	1.28E+02	20.000 ug/mL	
7 DUP	S98T002810 0	NH3-01	LIQUID	1.28E+2	1.62E+2	23.448 RPD	
8 STD	0	NH3-01	LIQUID	4.33E+02	3.64E+2	84.065 % Recovery	

Final page for worklist# 27849

Analyst Signature _____ Date _____

Analyst Signature *[Signature]* Date 3-15-99

Reviewer Signature *[Signature]* Date 3/15/99

Sample S98T002811
and S98T002810 were
Rejected Due to QC
Failures. *[Signature]*

DATA
NOT USED

01/11/99 10:59
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27849

Analyst: MF Instrument: NH301 Book# 50N19-11Method: LA-631-001 Rev/Mod D-0

Worklist Comment: BY 105 NH3-01 USE 1-10 DILUTION WITH 1ML SS MF

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			NH3-01	LIQUID		
2 STD			NH3-01	LIQUID		
3 SAMPLE	S98T002811 0		NH3-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T002811 0		NH3-01	LIQUID		
5 SPK	S98T002811 0		NH3-01	LIQUID		
6 SAM	S98T002810					
7 DUP						
8 STD			NH3-01	LIQUID		

Final page for worklist # 27849

Analyst Signature _____ Date _____

 Analyst Signature

 3-12-99
 Date
DATA
NOT USED

Data Entry Comments:

Add 598T002810 sample and dup

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED
AT 08:04, 03-12-99

HNF-1663 REV. 0

SAMPLE VOL= 25.000 AT 08:05, 03-12-99
ENTERED

MF= 194.6 mV AT 08:05, 03-12-99

MF= 191.3 mV AT 08:08, 03-12-99

MF= 189.6 mV AT 08:09, 03-12-99

MF= 189.1 mV AT 08:10, 03-12-99

MF= 188.7 mV AT 08:10, 03-12-99

MF= 188.4 mV AT 08:11, 03-12-99

MF= 187.7 mV AT 08:11, 03-12-99

MF= 187.5 mV AT 08:12, 03-12-99

MF= 187.1 mV AT 08:12, 03-12-99

MF= 186.7 mV AT 08:12, 03-12-99

MF= 186.4 mV AT 08:13, 03-12-99

MF= 186.2 mV AT 08:13, 03-12-99

MF= 185.9 mV AT 08:13, 03-12-99

MF= 185.8 mV AT 08:13, 03-12-99

ENTERED

STD CONC= 998 AT 08:14, 03-12-99
ENTERED

STD VOL= .25000 AT 08:14, 03-12-99
ENTERED

MF=-40.7 mV AT 08:19, 03-12-99

MF=-41.0 mV AT 08:19, 03-12-99

MF=-41.1 mV AT 08:19, 03-12-99

ENTERED

STD VOL= 2.5000 AT 08:20, 03-12-99
ENTERED

MF=-101.4 mV AT 08:23, 03-12-99

MF=-101.7 mV AT 08:23, 03-12-99

MF=-102.0 mV AT 08:23, 03-12-99

MF=-102.1 mV AT 08:24, 03-12-99

ENTERED

..NH3 SLOPE=-61.0 mV/DEC
AT 08:24, 03-12-99

..NH3 CONC= .00188

BLK

964

DATA
NOT USED

DOUBLE KNOWN ADDITION SELECTED
AT 08:39, 03-12-99

SAMPLE VOL= 25.000 AT 08:40, 03-12-99
ENTERED

EMF=-32.7 mV AT 08:43, 03-12-99

EMF=-33.0 mV AT 08:43, 03-12-99

EMF=-33.2 mV AT 08:43, 03-12-99
ENTERED

STD CONC= 998 AT 08:44, 03-12-99
ENTERED

STD VOL= .25000 AT 08:44, 03-12-99
ENTERED

EMF=-51.7 mV AT 08:46, 03-12-99

EMF=-52.3 mV AT 08:46, 03-12-99

EMF=-52.1 mV AT 08:47, 03-12-99

EMF=-52.7 mV AT 08:47, 03-12-99

EMF=-52.6 mV AT 08:47, 03-12-99
ENTERED

STD VOL= 2.5000 AT 08:48, 03-12-99
ENTERED

EMF=-97.0 mV AT 08:50, 03-12-99

EMF=-97.2 mV AT 08:51, 03-12-99

EMF=-97.4 mV AT 08:51, 03-12-99

EMF=-97.4 mV AT 08:52, 03-12-99
ENTERED

..NH3 SLOPE=-58.7 mV/DEC
AT 08:52, 03-12-99

..NH3 CONC= 8.58

.500 ml 50N19-H

DATA
NOT USED

DOUBLE KNOWN ADDITION SELECTED
AT 09:03, 03-12-99

SAMPLE VOL= 25.000 AT 09:03, 03-12-99
ENTERED

MF=-4.7 mV AT 09:06, 03-12-99

MF=-4.9 mV AT 09:06, 03-12-99

MF=-5.2 mV AT 09:06, 03-12-99

MF=-5.3 mV AT 09:07, 03-12-99

MF=-5.6 mV AT 09:07, 03-12-99

EMF=-5.7 mV AT 09:07, 03-12-99

EMF=-5.7 mV AT 09:08, 03-12-99

EMF=-5.7 mV AT 09:08, 03-12-99

EMF=-5.7 mV AT 09:09, 03-12-99

EMF=-5.7 mV AT 09:09, 03-12-99

ENTERED

STD CONC= 998 AT 09:09, 03-12-99
ENTERED

STD VOL= .25000 AT 09:09, 03-12-99
ENTERED

EMF=-42.7 mV AT 09:12, 03-12-99

EMF=-43.0 mV AT 09:12, 03-12-99

EMF=-43.1 mV AT 09:13, 03-12-99

ENTERED

STD VOL= 2.5000 AT 09:13, 03-12-99
ENTERED

EMF=-94.2 mV AT 09:16, 03-12-99

EMF=-94.5 mV AT 09:16, 03-12-99

EMF=-94.5 mV AT 09:16, 03-12-99

ENTERED

-.NH3 SLOPE=-56.8 mV/DEC
AT 09:16, 03-12-99

-.NH3 CONC= 2.77 .250 ml 2811

DOUBLE KNOWN ADDITION SELECTED
AT 09:17, 03-12-99

DATA
NOT USED

AMPLE VOL= 25.000 AT 10:55, 03-12-99
ENTERED

EMF= 27.2 mV AT 11:00, 03-12-99

EMF= 26.8 mV AT 11:00, 03-12-99

EMF= 26.5 mV AT 11:00, 03-12-99

EMF= 25.9 mV AT 11:01, 03-12-99

EMF= 25.6 mV AT 11:02, 03-12-99
ENTERED

STD CONC= 998 AT 11:02, 03-12-99
ENTERED

STD VOL= .25000 AT 11:02, 03-12-99
ENTERED

EMF=-12.9 mV AT 11:07, 03-12-99

EMF=-13.3 mV AT 11:07, 03-12-99

EMF=-13.4 mV AT 11:07, 03-12-99
ENTERED

STD VOL= 2.5000 AT 11:08, 03-12-99
ENTERED

EMF=-69.6 mV AT 11:12, 03-12-99

EMF=-70.1 mV AT 11:12, 03-12-99

EMF=-70.3 mV AT 11:13, 03-12-99
ENTERED

NH3 SLOPE=-63.8 mV/DEC
AT 11:13, 03-12-99

NH3 CONC= 3.19

.250ml 2811 Dup

DATA
NOT USED

AMPLE VOL= 25.000 AT 10:32, 03-12-99
ENTERED

HNF-1663 REV. 0

MF= 23.7 mV AT 10:37, 03-12-99

MF= 22.6 mV AT 10:38, 03-12-99

MF= 22.3 mV AT 10:38, 03-12-99

MF= 21.9 mV AT 10:39, 03-12-99

MF= 21.6 mV AT 10:39, 03-12-99
ENTERED

STD CONC= 998 AT 10:39, 03-12-99
ENTERED

STD VOL= .25000 AT 10:39, 03-12-99
ENTERED

EMF=-0.7 mV AT 10:41, 03-12-99

EMF=-3.3 mV AT 10:42, 03-12-99

EMF=-4.1 mV AT 10:43, 03-12-99

EMF=-4.4 mV AT 10:43, 03-12-99

EMF=-4.9 mV AT 10:43, 03-12-99

EMF=-5.0 mV AT 10:44, 03-12-99
ENTERED

STD VOL= 2.5000 AT 10:44, 03-12-99
ENTERED

EMF=-53.8 mV AT 10:47, 03-12-99

EMF=-54.4 mV AT 10:48, 03-12-99

EMF=-54.8 mV AT 10:48, 03-12-99

EMF=-54.9 mV AT 10:48, 03-12-99
ENTERED

DATA
NOT USED

NH3 SLOPE=-60.1 mV/DEC
AT 10:48, 03-12-99

NH3 CONC= 5.56 .250ml 2811 SPK + .250ml 50N19-H

DOUBLE KNOWN ADDITION SELECTED
AT 10:48, 03-12-99

DOUBLE KNOWN ADDITION SELECTED
AT 11:27, 03-12-99

SAMPLE VOL= 25.000 AT 11:27, 03-12-99
ENTERED

EMF= 38.8 mV AT 11:29, 03-12-99

EMF= 38.6 mV AT 11:30, 03-12-99

EMF= 38.6 mV AT 11:30, 03-12-99
ENTERED

STD CONC= 998 AT 11:30, 03-12-99
ENTERED

STD VOL= .25000 AT 11:30, 03-12-99
ENTERED

EMF=-16.8 mV AT 11:33, 03-12-99

EMF=-17.2 mV AT 11:33, 03-12-99

EMF=-17.5 mV AT 11:33, 03-12-99

EMF=-17.5 mV AT 11:33, 03-12-99
ENTERED

STD VOL= 2.5000 AT 11:34, 03-12-99
ENTERED

EMF=-73.8 mV AT 11:37, 03-12-99

EMF=-74.1 mV AT 11:37, 03-12-99

EMF=-74.4 mV AT 11:37, 03-12-99

EMF=-74.4 mV AT 11:37, 03-12-99
ENTERED

NH3 SLOPE=-59.7 mV/DEC
AT 11:37, 03-12-99

NH3 CONC= 1.28   *.250ml 2810*

DATA
NOT USED

DOUBLE KNOWN ADDITION SELECTED

AT 11:47, 03-12-99

SAMPLE VOL= 25.000 AT 11:47, 03-12-99
ENTERED

EMF= 32.3 mV AT 11:49, 03-12-99

EMF= 32.1 mV AT 11:49, 03-12-99

EMF= 31.9 mV AT 11:50, 03-12-99
ENTEREDSTD CONC= 998 AT 11:50, 03-12-99
ENTEREDSTD VOL= .25000 AT 11:50, 03-12-99
ENTERED

EMF=-18.3 mV AT 11:53, 03-12-99

EMF=-18.6 mV AT 11:54, 03-12-99

EMF=-18.9 mV AT 11:54, 03-12-99

EMF=-19.0 mV AT 11:54, 03-12-99
ENTEREDSTD VOL= 2.5000 AT 11:56, 03-12-99
ENTERED

EMF=-74.4 mV AT 11:58, 03-12-99

EMF=-74.8 mV AT 11:59, 03-12-99

EMF=-75.2 mV AT 11:59, 03-12-99

EMF=-75.4 mV AT 11:59, 03-12-99
ENTERED1:NH3 SLOPE=-59.9 mV/DEC
AT 11:59, 03-12-99

1:NH3 CONC= 1.62 .250 ml 2810 Dup

DATA
NOT USED

DOUBLE KNOWN ADDITION SELECTED
AT 13:22, 03-12-99

SAMPLE VOL= 25.000 AT 13:22, 03-12-99
ENTERED

MF= 33.7 mV AT 13:26, 03-12-99

MF= 32.7 mV AT 13:27, 03-12-99

MF= 30.7 mV AT 13:28, 03-12-99

MF= 30.1 mV AT 13:29, 03-12-99

MF= 29.3 mV AT 13:30, 03-12-99

MF= 29.1 mV AT 13:30, 03-12-99

ENTERED

STD CONC= 998 AT 13:30, 03-12-99
ENTERED

STD VOL= .25000 AT 13:30, 03-12-99
ENTERED

MF= 8.1 mV AT 13:32, 03-12-99

MF= 7.0 mV AT 13:33, 03-12-99

MF= 6.8 mV AT 13:33, 03-12-99

MF= 6.8 mV AT 13:33, 03-12-99
ENTERED

STD VOL= 2.5000 AT 13:34, 03-12-99
ENTERED

MF=-39.4 mV AT 13:37, 03-12-99

MF=-40.0 mV AT 13:38, 03-12-99

MF=-40.5 mV AT 13:38, 03-12-99

MF=-40.8 mV AT 13:39, 03-12-99
ENTERED

NH3 SLOPE=-60.2 mV/DEC
AT 13:39, 03-12-99

NH3 CONC= 7.29

500 ml 50 N 19-H

DATA
NOT USED

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 ^D~~(E-0)~~ ⁴³₃₋₂₅₋₉₉ LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000097		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 2.00E+01
Sample #		
BLNK	Detection Limit (µg/mL)	2.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/15/99		
Analysis Date		
03/12/99		
Analysis Time		
02:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 2.00E+01
BY 105		

DATA
NOT USED

Analyst:	MF Date: 03/15/99
Signature of Chemist: <i>N/A</i>	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: STD2

AMMONIA (NH3) : LA-631-001 (E-0) ^D ₃₋₂₅₋₉₉

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (µg/mL)	ID 8.580
STD	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL 4.33E+02
Matrix		
LIQUID		
Batch Number		
99000097		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 4.29E+02
Sample #		
STD	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)	
MF		
Date Complete		
03/15/99		
Analysis Date	QC ACTUAL (µg)	4.33E+02
03/12/99	QC FOUND (µg)	4.29E+02
Analysis Time		
02:00 PM		
Sample Point		
BY 105		

DATA
NOT USED

Analyst:	MF	Date:	03/15/99
Signature of Chemist:	<i>NA</i>	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001

D
C-0 3-25-99
LAD

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 2.770
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000097		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.77E+02
Sample #		
S98T002811	Detection Limit (µg/mL)	2.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/15/99		
Analysis Date		
03/12/99		
Analysis Time		
02:00 PM		
Sample Point	NH3 Concentration (µg/mL)	2.77E+02
BY 105		

DATA
NOT USED

Analyst:	MF Date: 03/15/99
Signature of Chemist:	MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

AMMONIA (NH3) : LA-631-001 (E-0) ³⁻²⁵⁻⁹⁹₁₄₉

LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 3.190
DUP	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000097		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.19E+02
Sample #		
S98T002811	Detection Limit (µg/mL)	2.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/15/99		
Analysis Date		
03/12/99		
Analysis Time		
02:00 PM		
Sample Point	NH3 Concentration (µg/mL)	3.19E+02
BY 105		

DATA
NOT USED

Analyst:	MF	Date: 03/15/99
Signature of Chemist:	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SPIKE5

AMMONIA (NH3) : LA-631-001 (E-0) ^{MD} 3-25-99

LIQUIDS/SOLIDS

		SPK
Type	Instrument Data (µg/mL)	ID 5.560
SPK	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	Spike Book Number	SPK 50N19H
NH3-01	Spike Value (µg/mL)	SPK VAL 4.33E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK 0.250
LIQUID		
Batch Number		
99000097		
Rerun	Sample Instrument Data (µg/mL)	SAM ID 2.77
0	Sample Volume of Sample (mL)	SAM VSAM 0.250
Sample Prep	Sample Final Volume (mL)	SAM FVOL 25.0
N/A		
Sample #		
S98T002811		
Instrument Code		
NH302		
Prepared By		
JMV		
Chemist		
MJL	QC ACTUAL = SPK VAL * VOL SPK	
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))	
MF		
Date Complete		
03/15/99		
Analysis Date	QC ACTUAL (µg)	1.08E+02
03/12/99	QC FOUND (µg)	6.97E+01
Analysis Time		
02:00 PM		
Sample Point		
BY 105		

DATA
NOT USED

Analyst:	MF Date: 03/15/99
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

AMMONIA (NH3) : LA-631-001 (C-0)

LIQUIDS/SOLIDS

SAMPLE

Type	Instrument Data (µg/mL)	ID	1.280
SAMPLE	Blank Result from the Instrument (µg/mL)	BR	0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code			
NH3-01			
Matrix			
LIQUID	Dilution Factor	DF	1.000
Batch Number			
99000097			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	1.28E+02
Sample #			
S98T002810	Detection Limit (µg/mL)		2.00E+01
Instrument Code			
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF		
Prepared By			
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF		
Chemist			
MJL			
Analyst			
MF			
Date Complete			
03/15/99			
Analysis Date			
03/12/99			
Analysis Time			
02:00 PM			
Sample Point	NH3 Concentration (µg/mL)		1.28E+02
BY 105			

DATA
NOT USED

Analyst:	MF	Date:	03/15/99
Signature of Chemist:	NA	MJL	Date:

SAMPLE.WB1 REV 1.0

631001ML

AMMONIA (NH3) : LA-631-001 (C-0) ^D ^{14D} ³²⁵⁻⁴¹

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	DUP
DUP	Blank Result from the Instrument (µg/mL)	BR	0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.250
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code			
NH3-01			
Matrix			
LIQUID	Dilution Factor	DF	1.000
Batch Number			
99000097			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	1.62E+02
Sample #			
S98T002810	Detection Limit (µg/mL)	2.00E+01	
Instrument Code			
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF		
Prepared By			
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF		
Chemist			
MJL			
Analyst			
MF			
Date Complete			
03/15/99			
Analysis Date			
03/12/99			
Analysis Time			
02:00 PM			
Sample Point	NH3 Concentration (µg/mL)		1.62E+02
BY 105			

DATA
NOT USED

Analyst:	MF	Date:	03/15/99
Signature of Chemist:	<i>MF</i>	MJL	Date:
SAMPLE.WB1 REV 1.0			
631001ML			

WORKBOOK PAGE: ST_END8

AMMONIA (NH3) : LA-631-001 (E-0) ^{LD} 32549

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.500
27849	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	4.33E+02
Matrix			
Batch Number			
99000097			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	3.64E+02
Sample #			
	Detection Limit (µg/mL)		1.00E+01
Instrument Code			
NH302	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$		
Prepared By			
JMV	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
MF			
Date Complete			
03/15/99			
Analysis Date	QC ACTUAL (µg)		4.33E+02
03/12/99	QC FOUND (µg)		3.64E+02
Analysis Time			
02:00 PM			
Sample Point			
BY 105			

DATA
NOT USED

Analyst:	MF	Date:	03/15/99
Signature of Chemist:	<i>MA</i>	MJL	Date:
SAMPLE.WB1 REV 1.0		631001ML	

LABCORE Completed Worklist Report for Worklist# 28068

Analyst: mf Instrument: NH302 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY105 FOR NH3-01 RLC

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	BLNK	0		NH3-01	LIQUID	1	<1.00E+1		ug/mL
2	STD	0		NH3-01	LIQUID	4.33E+02	4.32E+2	99.769 % Recovery	
3	SAMPLE	S98T002391	0	NH3-01	LIQUID	N/A	2.97E+01	10.000	ug/mL
4	DUP	S98T002391	0	NH3-01	LIQUID	2.97E+1	3.81E+1	24.779 RPD	
5	STD	0		NH3-01	LIQUID	4.33E+02	3.93E+2	90.762 % Recovery	

Final page for worklist# 28068

Analyst Signature _____ Date _____

Leonidas 3-12-99
Analyst Signature _____ Date _____

[Signature] 3/15/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28068

Analyst: _____ Instrument: NH301 _____ Book# 50019-H

Method: LA-631-001 Rev/Mod D-0

Worklist Comment: BY105 FOR NH3-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			NH3-01	LIQUID		
2 STD			NH3-01	LIQUID		
3 SAMPLE	S98T002391 0		NH3-01	LIQUID	98000436	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T002391 0		NH3-01	LIQUID		
5 STD			NH3-01	LIQUID		

Final page for worklist # 28068

Analyst Signature _____ Date _____

Mary Jones 3-11-99/1500
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DOUBLE KNOWN ADDITION SELECTED

AT 08:15, 03-11-99

-AMPLE VOL= 25.000 AT 08:15, 03-11-99
ENTERED

MF= 194.0 mV AT 08:16, 03-11-99

MF= 193.7 mV AT 08:16, 03-11-99

MF= 192.9 mV AT 08:17, 03-11-99

MF= 192.6 mV AT 08:17, 03-11-99

MF= 192.2 mV AT 08:18, 03-11-99

IMF= 192.0 mV AT 08:18, 03-11-99

IMF= 191.6 mV AT 08:18, 03-11-99

IMF= 191.3 mV AT 08:19, 03-11-99

IMF= 191.0 mV AT 08:19, 03-11-99

EMF= 190.7 mV AT 08:20, 03-11-99

EMF= 190.4 mV AT 08:20, 03-11-99

EMF= 190.1 mV AT 08:20, 03-11-99

ENTERED

STD CONC= 998 AT 08:21, 03-11-99

ENTERED

STD VOL= .25000 AT 08:21, 03-11-99

ENTERED

IMF=-39.2 mV AT 08:25, 03-11-99

EMF=-39.5 mV AT 08:25, 03-11-99

IMF=-39.7 mV AT 08:25, 03-11-99

EMF=-40.1 mV AT 08:26, 03-11-99

EMF=-40.1 mV AT 08:26, 03-11-99

ENTERED

STD VOL= 2.5000 AT 08:27, 03-11-99

ENTERED

EMF=-99.2 mV AT 08:29, 03-11-99

EMF=-99.5 mV AT 08:29, 03-11-99

EMF=-99.6 mV AT 08:30, 03-11-99

ENTERED

2:INH3 SLOPE=-99.5 mV/DEC

AT 08:30, 03-11-99

1:INH3 CONC= .00133

BLK

DOUBLE KNOWN ADDITION SELECTED
AT 08:51, 03-11-99

SAMPLE VOL= 25.000 AT 08:51, 03-11-99
ENTERED

EMF=-35.6 mV AT 08:54, 03-11-99

EMF=-35.9 mV AT 08:55, 03-11-99

EMF=-36.0 mV AT 08:55, 03-11-99
ENTERED

STD CONC= 998 AT 08:55, 03-11-99
ENTERED

STD VOL= 25000 AT 08:56, 03-11-99
ENTERED

EMF=-54.3 mV AT 08:58, 03-11-99

EMF=-54.9 mV AT 08:58, 03-11-99
ENTERED

STD VOL= 2.5000 AT 08:59, 03-11-99
ENTERED

EMF=-98.3 mV AT 09:01, 03-11-99

EMF=-98.6 mV AT 09:01, 03-11-99

EMF=-98.7 mV AT 09:01, 03-11-99
ENTERED

L:NH3 SLOPE=-57.5 mV/DEC
AT 09:01, 03-11-99

L:NH3 CONC= 8.65

.500 ml 50N19-H

DOUBLE KNOWN ADDITION SELECTED
 AT 13:17, 03-11-99

SAMPLE VOL= 25.000 AT 13:17, 03-11-99
 ENTERED

EMF= 73.2 mV AT 13:23, 03-11-99

EMF= 72.7 mV AT 13:23, 03-11-99

EMF= 72.5 mV AT 13:23, 03-11-99
 ENTERED

STD CONCN= 998 AT 13:24, 03-11-99
 ENTERED

STD VOL= .25000 AT 13:24, 03-11-99
 ENTERED

EMF=-2.6 mV AT 13:27, 03-11-99

EMF=-2.5 mV AT 13:28, 03-11-99

EMF=-2.8 mV AT 13:28, 03-11-99

EMF=-3.2 mV AT 13:29, 03-11-99

EMF=-3.4 mV AT 13:29, 03-11-99
 ENTERED

STD VOL= 2.5000 AT 13:29, 03-11-99
 ENTERED

EMF=-62.3 mV AT 13:32, 03-11-99

EMF=-62.7 mV AT 13:33, 03-11-99

EMF=-63.0 mV AT 13:33, 03-11-99
 ENTERED

INHX SLOPE=-61.0 mV/DEC
 AT 13:33, 03-11-99

INHX CONCN= .596   .500 ml 2391

CUBLE KNOWN ADDITION SELECTED
AT 13:51, 03-11-99

AMPLE VOL= 25.000 AT 13:51, 03-11-99
ENTERED

MF= 65.9 mV AT 13:54, 03-11-99

IMF= 65.7 mV AT 13:54, 03-11-99

MF= 65.4 mV AT 13:55, 03-11-99

IMF= 65.4 mV AT 13:55, 03-11-99
ENTERED

STD CONC= .000 AT 13:56, 03-11-99
ENTERED

STD VOL= 1.5000 AT 13:56, 03-11-99
ENTERED

MF= 62.2 mV AT 13:59, 03-11-99

EMF=-2.6 mV AT 14:00, 03-11-99

EMF=-2.9 mV AT 14:00, 03-11-99

EMF=-3.2 mV AT 14:00, 03-11-99

EMF=-3.4 mV AT 14:01, 03-11-99

EMF=-3.6 mV AT 14:01, 03-11-99
ENTERED

STD VOL= 2.5000 AT 14:02, 03-11-99
ENTERED

EMF=-59.9 mV AT 14:04, 03-11-99

EMF=-60.9 mV AT 14:05, 03-11-99

EMF=-61.8 mV AT 14:05, 03-11-99

EMF=-62.1 mV AT 14:06, 03-11-99

EMF=-62.2 mV AT 14:06, 03-11-99
ENTERED

1: NH3 SLOPE=-60.3 mV/DEC
AT 14:06, 03-11-99

1: NH3 CONC= .764

.500 2391 Dup

DOUBLE KNOWN ADDITION SELECTED
AT 14:41, 03-11-99

AMPLE VOL= 25.000 AT 14:41, 03-11-99
ENTERED

MF=-19.7 mV AT 14:44, 03-11-99

IMF=-20.2 mV AT 14:44, 03-11-99

MF=-20.5 mV AT 14:45, 03-11-99

IMF=-20.6 mV AT 14:45, 03-11-99

MF=-20.8 mV AT 14:45, 03-11-99

IMF=-20.8 mV AT 14:45, 03-11-99
ENTERED

STD CONC= 998 AT 14:46, 03-11-99
ENTERED

STD VOL= .25000 AT 14:46, 03-11-99
ENTERED

EMF=-40.9 mV AT 14:48, 03-11-99

EMF=-41.1 mV AT 14:48, 03-11-99

EMF=-41.2 mV AT 14:49, 03-11-99

EMF=-41.4 mV AT 14:49, 03-11-99
ENTERED

STD VOL= 2.5000 AT 14:49, 03-11-99
ENTERED

EMF=-86.1 mV AT 14:52, 03-11-99

EMF=-86.4 mV AT 14:52, 03-11-99

EMF=-86.6 mV AT 14:53, 03-11-99

EMF=-86.7 mV AT 14:53, 03-11-99

EMF=-86.9 mV AT 14:53, 03-11-99

EMF=-87.0 mV AT 14:54, 03-11-99
ENTERED

1: NH3 SLOPE=-58.6 mV/DEC
AT 14:54, 03-11-99

1: NH3 CONC= 7.87

.500 ml 50N19-H

WORKBOOK PAGE: BLANK1

AMMONIA (NH3) : LA-631-001 (C-0)

LAD
3-25-99

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.001
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
28068	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000344		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 1.00E+01
Sample #		
BLNK	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/12/99		
Analysis Date		
03/11/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 1.00E+01
BY 105		

Analyst:

MF Date: 03/12/99

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

AMMONIA (NH3) : LA-631-001 (Q-0) 3-28-99

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	0.001
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.500
28068	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	4.33E+02
Matrix			
Batch Number			
99000344			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	4.32E+02
Sample #			
STD	Detection Limit (µg/mL)	1.00E+01	
Instrument Code			
NH302	Detection Limit = 0.200µg * (FVOL/VSAM)		
Prepared By			
JMV	NH3 Concentration (µg/mL) = (ID-BR)*(FVOL /VSAM)		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = (ID-BR) * (FVOL/VSAM)		
MF			
Date Complete			
03/12/99			
Analysis Date	QC ACTUAL (µg)		4.33E+02
03/11/99	QC FOUND (µg)		4.32E+02
Analysis Time			
03:00 PM			
Sample Point			
BY 105			

Analyst:

MF Date: 03/12/99

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (D-0) 3254

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.596
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.001
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
28068	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000344		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.97E+01
Sample #		
S98T002391	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/12/99		
Analysis Date		
03/11/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	2.97E+01
BY 105		

Analyst:

MF Date: 03/12/99

Signature of Chemist:

MJL Date:

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (E-0) ^D LAP 325-75 LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 0.764
DUP	Blank Result from the Instrument (µg/mL)	BR 0.001
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
28068	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99000344		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 3.81E+01
Sample #		
S98T002391	Detection Limit (µg/mL)	1.00E+01
Instrument Code		
NH302	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
JMV	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
03/12/99		
Analysis Date		
03/11/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	3.81E+01
BY 105		

Analyst:	MF Date: 03/12/99
Signature of Chemist:	MJL Date:
SAMPLE.WB1 REV 1.0	631001ML

WORKBOOK PAGE: ST_END5

AMMONIA (NH3) : LA-631-001 (C-0) 40325-99 LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	0.001
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.500
28068	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	4.33E+02
Matrix			
Batch Number			
99000344			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	3.93E+02
Sample #			
	Detection Limit (µg/mL)		1.00E+01
Instrument Code			
NH302	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$		
Prepared By			
JMV	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
MF			
Date Complete			
03/12/99			
Analysis Date	QC ACTUAL (µg)		4.33E+02
03/11/99	QC FOUND (µg)		3.93E+02
Analysis Time			
03:00 PM			
Sample Point			
BY 105			

Analyst:	MF	Date:	03/12/99
Signature of Chemist:	<i>NA</i>	MJL	Date:
SAMPLE.WB1 REV 1.0	651001ML		

LABCORE Completed Worklist Report for Worklist# 28749

Analyst: mf Instrument: NH301 Book#: _____

Method: LA-631-001 Rev/Mod _____

Worklist Comment: BY ^{105 m/L 3/17/99} ~~101~~ RERUN NH301 JMV

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	NH3-01	LIQUID	1	<2.50E+1		ug/mL
2 STD	0	NH3-01	LIQUID	4.33E+02	4.20E+2	96.998 % Recovery	
3 SAMPLE	S98T002810 0	NH3-01	LIQUID	N/A	1.07E+02	25.000 ug/mL	
4 DUP	S98T002810 0	NH3-01	LIQUID	1.07E+2	1.37E+2	24.590 RPD	
5 SAMPLE	S98T002811 0	NH3-01	LIQUID	N/A	2.00E+02	25.000 ug/mL	
6 DUP	S98T002811 0	NH3-01	LIQUID	2.00E+2	1.57E+2	24.090 RPD	
7 SPK	S98T002811 0	NH3-01	LIQUID	1.08E+02	1.22E+02	112.963 % Recovery	
8 STD	0	NH3-01	LIQUID	4.33E+02	4.32E+2	99.769 % Recovery	

Final page for worklist# 28749

Analyst Signature _____ Date _____

Mary Innes 3-17-99
Analyst Signature Date

[Signature] 3/17/99
Reviewer Signature Date

m/L Sample rejected due to standard
out of control. m/L m/L 3/17/99

LABCORE Data Entry Template for Worklist# 28749

Analyst: MF Instrument: NH301 Book# 502519-14

Method: LA-631-001 Rev/Mod D-0

Worklist Comment: BY 105(2) THE 3/15/99
LOT RERUN NH301 JMV

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			NH3-01	LIQUID		
2 STD			NH3-01	LIQUID		
3 SAMPLE	S98T002810 0		NH3-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: NH3-01					
4 DUP	S98T002810 0		NH3-01	LIQUID		
5 SAMPLE	S98T002811 0		NH3-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: NH3-01					
6 DUP	S98T002811 0		NH3-01	LIQUID		
7 SPK	S98T002811 0		NH3-01	LIQUID		
8 STD			NH3-01	LIQUID		

Final page for worklist # 28749

Signature _____ Date _____

Mary Jane 3-16-99
Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

TABLE KNOWN ADDITION SELECTED
T 07:57, 03-16-99

AMPLE VOL= 25.000 AT 07:57, 03-16-99
ENTERED

MF= 203.2 mV AT 07:58, 03-16-99

MF= 197.7 mV AT 08:02, 03-16-99

MF= 196.7 mV AT 08:02, 03-16-99

MF= 195.9 mV AT 08:03, 03-16-99

MF= 195.8 mV AT 08:03, 03-16-99
ENTERED

TD CONC= 998 AT 08:04, 03-16-99
ENTERED

TD VOL= .25000 AT 08:04, 03-16-99
ENTERED

MF=-32.2 mV AT 08:07, 03-16-99

MF=-32.8 mV AT 08:08, 03-16-99

MF=-33.0 mV AT 08:08, 03-16-99

MF=-33.3 mV AT 08:09, 03-16-99
ENTERED

STD VOL= 2.5000 AT 08:09, 03-16-99
ENTERED

EMF=-93.3 mV AT 08:12, 03-16-99

EMF=-93.6 mV AT 08:12, 03-16-99

EMF=-94.0 mV AT 08:13, 03-16-99

EMF=-94.0 mV AT 08:13, 03-16-99
ENTERED

1:NH3 SLOPE=-60.7 mV/DEC
AT 08:13, 03-16-99

1:NH3 CONC= .00166 *SEE*

UBLE KNOWN ADDITION SELECTED

T 08:29, 03-16-99

AMPLE VOL= 25.000 AT 08:29, 03-16-99

ENTERED

MF=-24.2 mV AT 08:32, 03-16-99

MF=-24.9 mV AT 08:33, 03-16-99

MF=-25.4 mV AT 08:33, 03-16-99

MF=-25.8 mV AT 08:33, 03-16-99

MF=-25.3 mV AT 08:34, 03-16-99

MF=-25.3 mV AT 08:34, 03-16-99

ENTERED

TD CONC= 998 AT 08:34, 03-16-99

ENTERED

TD VOL= .25000 AT 08:35, 03-16-99

ENTERED

MF=-43.2 mV AT 08:36, 03-16-99

MF=-43.6 mV AT 08:36, 03-16-99

MF=-44.0 mV AT 08:37, 03-16-99

MF=-44.3 mV AT 08:37, 03-16-99

MF=-44.6 mV AT 08:38, 03-16-99

MF=-44.6 mV AT 08:38, 03-16-99

ENTERED

STD VOL= 2.5000 AT 08:39, 03-16-99

ENTERED

MF=-88.0 mV AT 08:41, 03-16-99

MF=-88.3 mV AT 08:41, 03-16-99

MF=-88.5 mV AT 08:42, 03-16-99

MF=-88.7 mV AT 08:42, 03-16-99

ENTERED

1:NH3 SLOPE=-57.5 mV/DEC

AT 08:42, 03-16-99

1:NH3 CONC= 8.40 .500 ml 50N19-H

DOUBLE KNOWN ADDITION SELECTED

AT 08:50, 03-16-99

AMPLE VOL= 25.000 AT 08:50, 03-16-99

ENTERED

MF= 32.4 mV AT 08:53, 03-16-99

MF= 31.9 mV AT 08:54, 03-16-99

MF= 31.9 mV AT 08:54, 03-16-99

ENTERED

TD CONC= 998 AT 08:55, 03-16-99

ENTERED

TD VOL= .25000 AT 08:55, 03-16-99

ENTERED

EMF=-31.8 mV AT 08:58, 03-16-99

EMF=-31.4 mV AT 08:58, 03-16-99

EMF=-31.2 mV AT 08:58, 03-16-99

EMF=-31.1 mV AT 08:58, 03-16-99

ENTERED

STD VOL= 2.5000 AT 08:59, 03-16-99

ENTERED

EMF=-85.9 mV AT 09:01, 03-16-99

EMF=-86.3 mV AT 09:02, 03-16-99

EMF=-86.7 mV AT 09:02, 03-16-99

EMF=-86.7 mV AT 09:02, 03-16-99

ENTERED

1:NH3 SLOPE=-57.4 mV/DEC

AT 09:02, 03-16-99

1:NH3 CONC= .859

2810 .200 ml

DOUBLE KNOWN ADDITION SELECTED

AT 09:06, 03-16-99

AMPLE VOL= 25.000 AT 09:06, 03-16-99
ENTERED

MF= 25.1 mV AT 09:09, 03-16-99

MF= 24.6 mV AT 09:09, 03-16-99

MF= 24.5 mV AT 09:10, 03-16-99
ENTEREDTD CONC= 998 AT 09:10, 03-16-99
ENTEREDTD VOL= .25000 AT 09:10, 03-16-99
ENTERED

EMF=-32.1 mV AT 09:13, 03-16-99

EMF=-32.4 mV AT 09:14, 03-16-99
ENTEREDSTD VOL= 2.5000 AT 09:14, 03-16-99
ENTERED

EMF=-85.8 mV AT 09:16, 03-16-99

EMF=-86.2 mV AT 09:17, 03-16-99

EMF=-86.6 mV AT 09:17, 03-16-99

EMF=-86.7 mV AT 09:17, 03-16-99

EMF=-87.0 mV AT 09:18, 03-16-99

EMF=-87.1 mV AT 09:18, 03-16-99
ENTERED1:NH3 SLOPE=-57.0 mV/DEC
AT 09:18, 03-16-99

1:NH3 CONC= 1.10

.200ml 2810 days

DOUBLE KNOWN ADDITION SELECTED
AT 10:14, 03-16-99

SAMPLE VOL= 25.000 AT 10:14, 03-16-99
ENTERED

MF= 87.8 mV AT 10:15, 03-16-99

MF= 74.6 mV AT 10:17, 03-16-99

MF= 72.9 mV AT 10:17, 03-16-99

MF= 69.7 mV AT 10:18, 03-16-99

MF= 68.5 mV AT 10:19, 03-16-99

MF= 67.9 mV AT 10:19, 03-16-99

MF= 66.3 mV AT 10:20, 03-16-99

MF= 65.8 mV AT 10:21, 03-16-99

MF= 65.5 mV AT 10:21, 03-16-99

MF= 65.3 mV AT 10:21, 03-16-99
ENTERED

STD CONC= 998 AT 10:22, 03-16-99
ENTERED

STD VOL= .25000 AT 10:22, 03-16-99
ENTERED

MF= 14.7 mV AT 10:23, 03-16-99

MF= 12.4 mV AT 10:25, 03-16-99

MF= 12.5 mV AT 10:25, 03-16-99
ENTERED

STD VOL= 2.5000 AT 10:26, 03-16-99
ENTERED

MF=-42.9 mV AT 10:27, 03-16-99

MF=-43.8 mV AT 10:28, 03-16-99

MF=-44.3 mV AT 10:28, 03-16-99

MF=-44.9 mV AT 10:29, 03-16-99

MF=-45.2 mV AT 10:29, 03-16-99

MF=-45.5 mV AT 10:30, 03-16-99

MF=-45.6 mV AT 10:30, 03-16-99
ENTERED

1:NH3 SLOPE=-61.7 mV/DEC
AT 10:30, 03-16-99

1:NH3 CONC= 1.60

.200 ml 2811

DOUBLE KNOWN ADDITION SELECTED

AT 12:56, 03-16-99

AMPLE VOL= 25.000 AT 12:56, 03-16-99
ENTERED

MF= 28.9 mV AT 13:07, 03-16-99

MF= 28.4 mV AT 13:07, 03-16-99

MF= 28.0 mV AT 13:08, 03-16-99

MF= 27.9 mV AT 13:08, 03-16-99
ENTEREDTD CONC= 998 AT 13:08, 03-16-99
ENTEREDTD VOL= .25000 AT 13:09, 03-16-99
ENTERED

MF=-31.0 mV AT 13:14, 03-16-99

MF=-31.4 mV AT 13:14, 03-16-99

MF=-31.6 mV AT 13:14, 03-16-99
ENTEREDSTD VOL= 2.5000 AT 13:15, 03-16-99
ENTERED

MF=-90.6 mV AT 13:19, 03-16-99

MF=-91.2 mV AT 13:20, 03-16-99

MF=-91.5 mV AT 13:20, 03-16-99
ENTERED1:NH3 SLOPE=-62.8 mV/DEC
AT 13:20, 03-16-99

1:NH3 CONC= 1.26 , 200ml 2811 dup

*UBLE KNOWN ADDITION SELECTED

T 13:26, 03-16-99

MPLE VOL= 25.000 AT 13:26, 03-16-99

ENTERED

MF=-19.8 mV AT 13:28, 03-16-99

MF=-20.1 mV AT 13:28, 03-16-99

MF=-20.2 mV AT 13:29, 03-16-99

ENTERED

TD CONC= 998 AT 13:29, 03-16-99

ENTERED

TD VOL= .25000 AT 13:29, 03-16-99

ENTERED

MF=-42.2 mV AT 13:31, 03-16-99

MF=-42.9 mV AT 13:31, 03-16-99

MF=-43.2 mV AT 13:32, 03-16-99

MF=-43.4 mV AT 13:32, 03-16-99

ENTERED

TD VOL= 2.5000 AT 13:33, 03-16-99

ENTERED

EMF=-89.3 mV AT 13:34, 03-16-99

EMF=-89.7 mV AT 13:35, 03-16-99

EMF=-90.1 mV AT 13:36, 03-16-99

EMF=-90.2 mV AT 13:36, 03-16-99

ENTERED

1:NH3 SLOPE=-57.9 mV/DEC

AT 13:36, 03-16-99

1:NH3 CONC= 6.48 .200ml 2811 Spk 5019-H .250ml

DOUBLE KNOWN ADDITION SELECTED

AT 14:45, 03-16-99

AMPLE VOL= 25.000 AT 14:45, 03-16-99

ENTERED

EMF=-47.6 mV AT 14:49, 03-16-99

EMF=-47.9 mV AT 14:50, 03-16-99

EMF=-47.9 mV AT 14:50, 03-16-99

ENTERED

STD CONC= 998 AT 14:50, 03-16-99

ENTERED

STD VOL= .25000 AT 14:50, 03-16-99

ENTERED

EMF=-67.0 mV AT 14:54, 03-16-99

EMF=-67.2 mV AT 14:55, 03-16-99

ENTERED

STD VOL= 2.5000 AT 14:55, 03-16-99

ENTERED

EMF=-111.3 mV AT 14:58, 03-16-99

EMF=-111.6 mV AT 14:59, 03-16-99

EMF=-111.7 mV AT 14:59, 03-16-99

EMF=-111.9 mV AT 15:00, 03-16-99

EMF=-111.9 mV AT 15:00, 03-16-99

ENTERED

1:NH3 SLOPE=-58.6 mV/DEC

AT 15:00, 03-16-99

1:NH3 CONC= 8.64

.500 ml 50N19-H

AMMONIA (NH3) : LA-631-001

D
Q-0³⁻²⁵⁻⁹⁹
LA

LIQUIDS/SOLIDS

		BLNK
Type	Instrument Data (µg/mL)	ID 0.000
BLNK	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC < 2.50E+01
Sample #		
BLNK	Detection Limit (µg/mL)	2.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM)	
Prepared By		
MF	NH3 Concentration (µg/mL) = (BR) * (FVOL / VSAM) * DDF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
01/17/99		
Analysis Date		
01/16/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	< 2.50E+01
BY105		

Analyst:	MF	Date: 01/17/99
Signature of Chemist:	MJL	Date: 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

1002

WORKBOOK PAGE: STD2

AMMONIA (NH3) : LA-631-001 (C-0) ^D ^{LAD} ³⁻²⁵⁻⁹⁹

LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	STD
STD	Blank Result from the Instrument (µg/mL)	BR	0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.500
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	LCS Standard Book Number	LCS	50N19H
NH3-01	LCS Standard Concentration (µg/mL)	STD VAL	4.33E+02
Matrix			
LIQUID			
Batch Number			
99001072			
Rerun			
0			
Sample Prep			
N/A	NH3 Concentration (µg/mL)	NH3 CONC	4.20E+02
Sample #			
STD	Detection Limit (µg/mL)		1.00E+01
Instrument Code			
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$		
Prepared By			
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
Chemist			
MJL	QC ACTUAL = STD VAL		
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$		
MF			
Date Complete			
01/17/99			
Analysis Date	QC ACTUAL (µg)		4.33E+02
01/16/99	QC FOUND (µg)		4.20E+02
Analysis Time			
03:00 PM			
Sample Point			
BY105			

Analyst:	MF	Date: 07/17/99
Signature of Chemist:	MJL	Date: 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

1003

WORKBOOK PAGE: SAM3

AMMONIA (NH3) : LA-631-001 (0-0) ^D ~~3-28-99~~ ^{LAD}

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 0.859
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.07E+02
Sample #		
S98T002810	Detection Limit (µg/mL)	2.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
01/17/99		
Analysis Date		
01/16/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.07E+02
BY105		

Analyst:

MF Date: 01/17/99

Signature of Chemist:

MJL Date: ^{LAD} 3-28-99

SAMPLE.WB1 REV 1.0

631001ML

1004

WORKBOOK PAGE: DUP4

AMMONIA (NH3) : LA-631-001 (C-0) ^D _{LAD 3-25-99} LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 1.100
DUP	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.37E+02
Sample #		
S98T002810	Detection Limit (µg/mL)	2.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
01/17/99		
Analysis Date		
01/16/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.37E+02
BY105		

Analyst:	MF	Date: 01/17/99
Signature of Chemist:	MJL	Date: LAD 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

1005

WORKBOOK PAGE: SAM5

AMMONIA (NH3) : LA-631-001 (C-0) ^D 3-25-99
LAD

LIQUIDS/SOLIDS

		SAMPLE
Type	Instrument Data (µg/mL)	ID 1.600
SAMPLE	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 2.00E+02
Sample #		
S98T002811	Detection Limit (µg/mL)	2.50E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM}) * \text{DF}$	
Prepared By		
MF	NH3 Concentration (µg/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM}) * \text{DF}$	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
01/17/99		
Analysis Date		
01/16/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	2.00E+02
BY105		

Analyst:	MF	Date: 01/17/99
Signature of Chemist:	MJL	Date: 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

WORKBOOK PAGE: DUP6

AMMONIA (NH3) : LA-631-001 ^D~~(E-0)~~ ^{LAD} 3-25-99 LIQUIDS/SOLIDS

		DUP
Type	Instrument Data (µg/mL)	ID 1.260
DUP	Blank Result from the Instrument (µg/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code		
NH3-01		
Matrix		
LIQUID	Dilution Factor	DF 1.000
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (µg/mL)	NH3 CONC 1.57E+02
Sample #		
S98T002811	Detection Limit (µg/mL)	2.50E+01
Instrument Code		
NH301	Detection Limit = 0.200µg * (FVOL/VSAM) * DF	
Prepared By		
MF	NH3 Concentration (µg/mL) = (ID-BR) * (FVOL / VSAM)*DF	
Chemist		
MJL		
Analyst		
MF		
Date Complete		
01/17/99		
Analysis Date		
01/16/99		
Analysis Time		
03:00 PM		
Sample Point	NH3 Concentration (µg/mL)	1.57E+02
BY105		

Analyst:	MF	Date: 01/17/99
Signature of Chemist:	MJL	Date: 1-25-99

SAMPLE.WB1 REV 1.0

631001ML

1007

WORKBOOK PAGE: SPIKE7

AMMONIA (NH3) : LA-631-001 (0-0) ^D 3-25-99 _{LAD} LIQUIDS/SOLIDS

Type	Instrument Data (µg/mL)	ID	SPK
SPK	Blank Result from the Instrument (µg/mL)	BR	0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM	0.200
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL	25.0
Test Code	Spike Book Number	SPK	50N19H
NH3-01	Spike Value (µg/mL)	SPK VAL	4.33E+02
Matrix	Vol of Spike Standard Used (mL)	VOL SPK	0.250
LIQUID			
Batch Number			
99001072			
Rerun	Sample Instrument Data (µg/mL)	SAM ID	1.6
0	Sample Volume of Sample (mL)	SAM VSAM	0.200
Sample Prep	Sample Final Volume (mL)	SAM FVOL	25.0
N/A			
Sample #			
S98T002811			
Instrument Code			
NH301			
Prepared By			
MF			
Chemist			
MJL	QC ACTUAL = SPK VAL * VOL SPK		
Analyst	QC FOUND = (((ID-BR) * FVOL) - (SAM ID - BR) * SAM FVOL * (VSAM / SAM VSAM))		
MF			
Date Complete			
01/17/99			
Analysis Date	QC ACTUAL (µg)		1.08E+02
01/16/99	QC FOUND (µg)		1.22E+02
Analysis Time			
03:00 PM			
Sample Point			
BY105			

Analyst:	MF	Date:	01/17/99
Signature of Chemist:	<i>MA</i>	MJL	Date: <i>LAD</i> 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

1008

WORKBOOK PAGE: ST_END8

AMMONIA (NH3) : LA-631-001 (C-0)

LAD
3-25-99

LIQUIDS/SOLIDS

		STD
Type	Instrument Data (ug/mL)	ID 8.640
STD	Blank Result from the Instrument (ug/mL)	BR 0.002
Work List	Vol of Sample for Dilution (mL) or Vol of Sample Direct (mL)	VSAM 0.500
28749	Final Vol of Dilution (mL) or Vol of Sample Direct (mL)	FVOL 25.0
Test Code	LCS Standard Book Number	LCS 50N19H
NH3-01	LCS Standard Concentration (ug/mL)	STD VAL 4.33E+02
Matrix		
Batch Number		
99001072		
Rerun		
0		
Sample Prep		
N/A	NH3 Concentration (ug/mL)	NH3 CONC 4.32E+02
Sample #	Detection Limit (ug/mL)	1.00E+01
Instrument Code		
NH301	Detection Limit = $0.200\mu\text{g} * (\text{FVOL}/\text{VSAM})$	
Prepared By		
MF	NH3 Concentration (ug/mL) = $(\text{ID}-\text{BR}) * (\text{FVOL} / \text{VSAM})$	
Chemist		
MJL	QC ACTUAL = STD VAL	
Analyst	QC FOUND = $(\text{ID}-\text{BR}) * (\text{FVOL}/\text{VSAM})$	
MF		
Date Complete		
01/17/99		
Analysis Date	QC ACTUAL (ug)	4.33E+02
01/16/99	QC FOUND (ug)	4.32E+02
Analysis Time		
03:00 PM		
Sample Point		
BY105		

Analyst:	MF	Date: 01/17/99
Signature of Chemist:	MJL	Date: 3-25-99

SAMPLE.WB1 REV 1.0

631001ML

1009

03/17/99

08:38:22

11/24/98 08:54

LABCORE Data Entry Template for Worklist# 26739

Analyst: 115 Instrument: SP02 Book # 49N19DMethod: LA-265-101 Rev/Mod C - 0

Worklist Comment: BY-105 FOR CRVI-01 RTS

GROUP	PROJECT	S TYPE	SAMPLE#	R A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1 CCV			CRVI-01	LIQUID	<u>5.06</u>	<u>5.16E0</u>	<u>N/A</u>	ug/mL
		2 CCB			CRVI-01	LIQUID	<u>1</u>	<u><0.258</u>	<u>N/A</u>	ug/mL
98000436	BY-105 (2)	3 SAMPLE	S98T002392 0		CRVI-01	LIQUID	<u>N/A</u>	<u><0.258</u>		ug/mL
~~98000436~~	~~BY-105 (2)~~	~~4 DUP~~	~~S98T002392 0~~		~~CRVI-01~~	~~LIQUID~~	~~<0.258~~	~~<0.258~~	~~N/A~~	~~ug/mL~~
98000437	BY-105 (2)	5 SAMPLE	S98T002822 0		CRVI-01	LIQUID	<u>N/A</u>	<u>1.34E+1</u>		ug/mL
98000437	BY-105 (2)	6 DUP	S98T002822 0		CRVI-01	LIQUID	<u>1.34E+1</u>	<u>1.35E+1</u>	<u>N/A</u>	ug/mL
98000437	BY-105 (2)	7 SAMPLE	S98T002823 0		CRVI-01	LIQUID	<u>N/A</u>	<u>1.44E+1</u>		ug/mL
98000437	BY-105 (2)	8 DUP	S98T002823 0		CRVI-01	LIQUID	<u>1.44E+1</u>	<u>1.45E+1</u>	<u>N/A</u>	ug/mL

Final page for worklist # 26739

Analyst Signature _____ Date _____

Analyst Signature C. J. Pinner Date 11/24/98Validated 11/24/98 Rachel

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26739

Analyst: MS Instrument: SP02 Book# 24919-D

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002392 0		CRVI-01	LIQUID	98000436	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T002392 0		CRVI-01	LIQUID		
5 SAMPLE	S98T002822 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
6 DUP	S98T002822 0		CRVI-01	LIQUID		
7 SAMPLE	S98T002823 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
8 DUP	S98T002823 0		CRVI-01	LIQUID		

Final page for worklist # 26739

Analyst Signature Jeff Soltwick Date 10-30-98

Analyst Signature _____ Date _____

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		CCV
Type	Sample Absorbance at 540 nm (abs)	0.387
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26739	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.16E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By		
PPB	Less Than Value = abs < Low Calibration Point	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
11/10/98		
Analysis Date	v RESULT v	CCV
10/30/98		
Analysis Time	Cr +6 µg/mL =	5.16E+00
02:50 PM		
Sample Point		
BY-105		

Analyst:

JIS

Date:

11/10/98

Signature of Chemist:

PPB

Date:

STANDARD.WB1 REV 1.1

265101

1012

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26739	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun	
0	Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$
Sample Prep	If the result is a < value (abs < LOW) :
N/A	Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$
Sample #	RPD = $([Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP] \cdot 100 /$
CCB	(AVERAGE OF THE SAM + DUP)

Instrument Code	
SP02	Less Than Value = abs < Low Calibration Point

Prepared By	
PPB	
Chemist	
PPB	
Analyst	
JIS	

Date Complete	The Sample Absorbance is < Detection Limit (the Low Calibration Poi
11/10/98	

Analysis Date	v RESULT v	BLANK
10/30/98		
Analysis Time	Cr +6 µg/mL =	<0.258
02:50 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	11/10/98
Signature of Chemist:		PPB	Date:

SAMPLE.WB1 REV 1.

265101

1013

WORKBOOK PAGE: SAM4

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.271	0.272
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26739	Dilution Factor (DF)	1	1
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005340	Cr +6 µg/mL =	1.34E+01	1.35E+01
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002822	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*1		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
PPB			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
11/10/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
10/30/98	Cr +6 µg/mL =	1.34E+01	1.35E+01
Analysis Time			
02:50 PM			
Sample Point			
BY:105	Recovery % Difference =	0.37%	

Analyst:	JIS	Date:	11/10/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.291	0.292
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26739	Dilution Factor (DF)	1	1
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005340	Cr +6 µg/mL =	1.44E+01	1.45E+01

Rerun	
0	$Cr +6 \mu g/mL = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$
Sample Prep	If the result is a < value (abs < LOW) :
N/A	$Cr +6 \mu g/mL = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$
Sample #	
S98T002823	$RPD = (Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP / AVERAGE OF THE SAM + DUP) \cdot 1$
Instrument Code	Less Than Value = abs < Low Calibration Point

SP02
Prepared By
PPB
Chemist
PPB
Analyst
JIS
Date Complete
11/10/98
Analysis Date
10/30/98
Analysis Time
02:50 PM
Sample Point
BY-105

v RESULT v	SAMPLE	DUPLICATE
Cr +6 µg/mL =	1.44E+01	1.45E+01
Recovery % Difference =	0.35%	

Analyst:	JIS	Date:	11/10/98
Signature of Chemist:		PPB	Date:

SAMPLE.WB1 REV 1.

265101

LABCORE Completed Worklist Report for Worklist# 26740

Analyst: jis Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCV	0		CRVI-01	LIQUID	5.06	5.08E+00	100.395 %	Recovery
2	CCB	0		CRVI-01	LIQUID	1	<2.58E-1		Ratio
3	SAMPLE	S98T002824	0	CRVI-01	LIQUID	N/A	1.40E+3	1.00E-002	ug/mL
4	DUP	S98T002824	0	CRVI-01	LIQUID	1.40E+3	1.40E+3	0.000	RPD
5	SAMPLE	S98T002825	0	CRVI-01	LIQUID	N/A	1.51E+3	1.00E-002	ug/mL
6	DUP	S98T002825	0	CRVI-01	LIQUID	1.51E+3	1.52E+3	0.660	RPD
7	SAMPLE	S98T002826	0	CRVI-01	LIQUID	N/A	1.53E+3	1.00E-002	ug/mL
8	DUP	S98T002826	0	CRVI-01	LIQUID	1.53E+3	1.55E+3	1.299	RPD

Final page for worklist# 26740

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

3/9/19
Reviewer Signature Date

11/11/98 09:53
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26740

Analyst: JS Instrument: SP02 _____ Book# 49119-8Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002824 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T002824 0		CRVI-01	LIQUID		
5 SAMPLE	S98T002825 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
6 DUP	S98T002825 0		CRVI-01	LIQUID		
7 SAMPLE	S98T002826 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
8 DUP	S98T002826 0		CRVI-01	LIQUID		

Final page for worklist # 26740

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE****CCV**

Type	Sample Absorbance at 540 nm (abs)	0.381
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26740	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.08E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By		
CJO	Less Than Value = abs < Low Calibration Point	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
11/25/98		
Analysis Date	v RESULT v	CCV
11/12/98		
Analysis Time	Cr +6 µg/mL =	5.08E+00
02:40 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	11/25/98
Signature of Chemist:	PPB	Date:	

STANDARD.WB1 REV 1.1

265101

1018

03/08/99

16:18:29

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26740	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun	
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)
Sample Prep	If the result is a < value (abs < LOW) :
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)
Sample #	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP * 100 /
BLANK	(AVERAGE OF THE SAM + DUP)

Instrument Code	
SP02	Less Than Value = abs < Low Calibration Point

Prepared By	
CJO	
Chemist	
PPB	
Analyst	
JIS	

Date Complete	The Sample Absorbance is < Detection Limit (the Low Calibration Poi
11/25/98	

Analysis Date	v RESULT v	BLANK
11/12/98		
Analysis Time	Cr +6 µg/mL =	<0.258
02:40 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	11/25/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:
SAMPLE.WB1 REV 1.	265101		

1019

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.280	0.280
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26740	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005341	Cr +6 µg/mL =	1.40E+03	1.40E+03

Rerun
0
Sample Prep
N/A
Sample #
S98T002824
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
11/25/98
Analysis Date
11/12/98
Analysis Time
02:40 PM
Sample Point
BY-105

Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$

If the result is a < value (abs < LOW) :

Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$ RPD = $|(Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP) / AVERAGE OF THE SAM + DUP|$

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL =

	SAMPLE	DUPLICATE
	1.40E+03	1.40E+03
Recovery % Difference =	0.00%	

Analyst:

JIS

Date: 11/25/98

Signature of Chemist:

PPB

Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM4

Cr+6 : LA-265-101 (C-0)

LIQUID/SOLID SAMPLE

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.301	0.303
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26740	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005341	Cr +6 µg/mL =	1.51E+03	1.52E+03

Rerun	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
0	If the result is a < value (abs < LOW) :		
Sample Prep	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
N/A			
Sample #	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
S98T002825	Less Than Value = abs < Low Calibration Point		
Instrument Code			
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
11/25/98			
Analysis Date	v RESULT v		
11/12/98			
Analysis Time	Cr +6 µg/mL =	1.51E+03	1.52E+03
02:40 PM			
Sample Point			
BY-105	Recovery % Difference =	0.67%	

Analyst:	JIS	Date:	11/25/98
Signature of Chemist:	PPB	Date:	
SAMPLE.WB1 REV 1.	265101		

1021

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)

LIQUID/SOLID SAMPLE

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.305	0.309
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26740	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005341	Cr +6 µg/mL =	1.53E+03	1.55E+03

Rerun
0
Sample Prep
N/A
Sample #
S98T002826
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
11/25/98
Analysis Date
11/12/98
Analysis Time
02:40 PM
Sample Point
BY-105

Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)

If the result is a < value (abs < LOW) :

Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)

RPD = (|Cr +6 µg/mL SAM - Cr +6 µg/mL DUP| / AVERAGE OF THE SAM + DUP)*

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL =

SAMPLE	DUPLICATE
1.53E+03	1.55E+03
Recovery % Difference = 1.31%	

Analyst:	JIS	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.

265101

1022

LABCORE Data Entry Template for Worklist# 26741

Analyst: JIS Instrument: SP02 Book # 49N19D

Method: LA-265-101 Rev/Mod C-D

Worklist Comment: BY-105 FOR CRVI-01 RTS

GROUP	PROJECT	S	TYPE	SAMPLE#	R	A	-----TEST-----	MATRIX	ACTUAL	FOUND	DL	UNIT
		1	CCV				CRVI-01	LIQUID	<u>5.06</u>	<u>5.06</u>	<u>N/A</u>	ug/mL
		2	CCB				CRVI-01	LIQUID	<u>1</u>	<u>20.258</u> <u>97.186</u>	<u>12/3/98</u> <u>N/A</u>	ug/mL
98000437	BY-105 (2)	3	SAMPLE	S98T002827	0		CRVI-01	LIQUID	<u>N/A</u>	<u>1.03E+3</u>		ug/mL
98000437	BY-105 (2)	4	DUP	S98T002827	0		CRVI-01	LIQUID	<u>1.03E+3</u>	<u>1.04E+3</u>	<u>N/A</u>	ug/mL
98000437	BY-105 (2)	5	SPK	S98T002827	0		CRVI-01	LIQUID	<u>100</u>	<u>98.0</u>	<u>N/A</u>	ug/mL
98000437	BY-105 (2)	6	SAMPLE	S98T002828	0		CRVI-01	LIQUID	<u>N/A</u>	<u>1.35E+3</u>		ug/mL
98000437	BY-105 (2)	7	DUP	S98T002828	0		CRVI-01	LIQUID	<u>1.35E+3</u>	<u>1.37E+3</u>	<u>N/A</u>	ug/mL
98000437	BY-105 (2)	8	SAMPLE	S98T002829	0		CRVI-01	LIQUID	<u>N/A</u>	<u>1.33E+3</u>		ug/mL
98000437	BY-105 (2)	9	DUP	S98T002829	0		CRVI-01	LIQUID	<u>1.33E+3</u>	<u>1.32E+3</u>	<u>N/A</u>	ug/mL

Final page for worklist # 26741

See Attached for Signature
Analyst Signature Date

Cf. Omin 12/1/98
Analyst Signature Date

Validated 12/3/98 [Signature]
Validated Date Analyst

Data Entry Comments:

Units shown for QC (SPK & STD) may not reflect the actual units. DL = Detection Limit, S = Worklist Slot Number,
R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26741

Analyst: JIS Instrument: SP02 Book# 4 N19-D

Method: LA-265-101 Rev/Mod C-O

Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002827 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T002827 0		CRVI-01	LIQUID		
5 SPK	S98T002827 0		CRVI-01	LIQUID		
6 SAMPLE	S98T002828 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
7 DUP	S98T002828 0		CRVI-01	LIQUID		
8 SAMPLE	S98T002829 0		CRVI-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CRVI-01					
9 DUP	S98T002829 0		CRVI-01	LIQUID		

Final page for worklist # 26741

Jeff Saltnack 11-13-98
Analyst Signature Date

Ch. Gann 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE****CCV**

Type	Sample Absorbance at 540 nm (abs)	0.380
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26741	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
crvi-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.06E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By		
CJO	Less Than Value = abs < Low Calibration Point	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
12/01/98		
Analysis Date	v RESULT v	
11/13/98		
Analysis Time	Cr +6 µg/mL =	5.06E+00
10:55 AM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	PPB	Date:

STANDARD.WB1 REV 1.1

265101

1025

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	BLANK
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26741	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
BLANK
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/01/98
Analysis Date
11/13/98
Analysis Time
10:55 AM
Sample Point
BY-105

Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$
 If the result is a < value (abs < LOW) :
 Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$
 RPD = $(|Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP| \cdot 100 /$
 (AVERAGE OF THE SAM + DUP)

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1026

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)

LIQUID/SOLID SAMPLE

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.206	0.209
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26741	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005342	Cr +6 µg/mL =	1.03E+03	1.04E+03
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002827	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
12/01/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
11/13/98			
Analysis Time	Cr +6 µg/mL =	1.03E+03	1.04E+03
10:55 AM			
Sample Point			
BY-105	Recovery % Difference =	1.47%	

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1027

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

	SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs) 0.206
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs) 0.55
Work List	SAMPLE VOLUME (Sample) in mL (Vsam) 0.250
26741	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam) 0.250
Test Code	VOLUME of SPIKE in mL (VSPK) 1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF) 10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF) 11.25
LIQUID	Sample Dilution Factor (DF) 101
Batch Number	Digest Dilution Factor (DDF)
98005342	SPIKE DILUTION FACTOR (SDF) 1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL) 5.07
0	
Sample Prep	LOW CALIBRATION POINT 0.022
N/A	HIGH CALIBRATION POINT 0.831
Sample #	Y-INTERCEPT (b) 0.00276
S98T002827	SLOPE (m) 0.8198
Instrument Code	
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)
CJO	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100
Chemist	
PPB	
Analyst	
JIS	
Date Complete	
12/01/98	
Analysis Date	
11/13/98	
Analysis Time	RESULT
10:55 AM	QC ACTUAL = 5.07E+00
Sample Point	QC FOUND = 4.97E+00
BY-105	% SPIKE RECOVERY = 98.0%

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	PPB	Date:	

SPIKE.WB1 REV 1.0

265101

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.270	0.273
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26741	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
crvi-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005342	Cr +6 µg/mL =	1.35E+03	1.37E+03

Rerun
0
Sample Prep
N/A
Sample #
S98T002828
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/01/98
Analysis Date
11/13/98
Analysis Time
10:55 AM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} \square < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|Cr +6 \mu\text{g/mL SAM} - Cr +6 \mu\text{g/mL DUP}| / \text{AVERAGE OF THE SAM} + \text{DUP}) \cdot 100$$

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL □

Recovery % Difference = 1.12%

SAMPLE	DUPLICATE
1.35E+03	1.37E+03

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM6

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.267	0.265
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26741	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005342	Cr +6 µg/mL =	1.33E+03	1.32E+03

Rerun
0
Sample Prep
N/A
Sample #
S98T002829
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/01/98
Analysis Date
11/13/98
Analysis Time
10:55 AM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|Cr +6 \mu\text{g/mL SAM} - Cr +6 \mu\text{g/mL DUP}| / \text{AVERAGE OF THE SAM + DUP}) \cdot 100$$

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL =

SAMPLE	DUPLICATE
1.33E+03	1.32E+03

Recovery % Difference = 0.76%

Analyst:	JIS	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

LABCORE Completed Worklist Report for Worklist# 26742

Analyst: jis

Instrument: SP02

Book#: 49N19D

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCV	0	CRVI-01	LIQUID	5.06	5.18	102.372 % Recovery	
2 CCB	0	CRVI-01	LIQUID	1	<0.258	Ratio	
3 SAMPLR	S98T002940 0	CRVI-01	LIQUID	N/A	9.41e+2	1.00e-002 ug/mL	
4 DUP	S98T002940 0	CRVI-01	LIQUID	9.41e2	9.31e+2	1.068 RPD	
5 SPK	S98T002940 0	CRVI-01	LIQUID	100	113.0	113.000 % Recovery	
6 SAMPLR	S98T002965 0	CRVI-01	LIQUID	N/A	1.12e+3	1.00e-002 ug/mL	
7 DUP	S98T002965 0	CRVI-01	LIQUID	1.12e+3	1.12e+3	0.000 RPD	
8 SPK	S98T002965 0	CRVI-01	LIQUID	100	99.3	99.300 % Recovery	
9 SAMPLR	S98T002966 0	CRVI-01	LIQUID	N/A	1.10e+3	1.00e-002 ug/mL	
10 DUP	S98T002966 0	CRVI-01	LIQUID	1.10e+3	1.09e+3	0.913 RPD	

Final page for worklist# 26742

Analyst Signature _____ Date _____

Analyst Signature *J. J. J.* Date 12/1/98

Reviewer Signature *B. Rachel* Date 12/2/98

LABCORE Data Entry Template for Worklist# 26742

Analyst: JIS Instrument: SP02 Book# 24919-D-005 11/10/98
Method: LA-265-101 Rev/Mod C-0 49N19D
Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002940 0		CRVI-01	LIQUID	98000490	BY-105 (2)
	Analytes Requested:		CRVI-01			
4 DUP	S98T002940 0		CRVI-01	LIQUID		
5 SPK	S98T002940 0		CRVI-01	LIQUID		
6 SAMPLE	S98T002965 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested:		CRVI-01			
7 DUP	S98T002965 0		CRVI-01	LIQUID		
8 SPK	S98T002965 0		CRVI-01	LIQUID		
9 SAMPLE	S98T002966 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested:		CRVI-01			
10 DUP	S98T002966 0		CRVI-01	LIQUID		

Final page for worklist # 26742

Jeff Goldbach 10-30-98
Analyst Signature Date

C.D. Opium 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	CCV
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26742	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.18E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SPO2	(AVERAGE OF THE SAM + DUP)	
Prepared By	Less Than Value = abs < Low Calibration Point	
PPB		
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
11/18/98		
Analysis Date	v RESULT v	CCV
10/30/98	Cr +6 µg/mL =	5.18E+00
Analysis Time		
12:50 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	11/18/98
Signature of Chemist:	<i>1/4</i>	PPB	Date:

STANDARD.WB1 REV 1.1

265101

1033

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	BLANK
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26742	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258
Rerun		
0	Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$	
Sample Prep	If the result is a < value (abs < LOW) :	
N/A	Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$	
Sample #	RPD = $(Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP \cdot 100 /$	
CCB	(AVERAGE OF THE SAM + DUP)	
Instrument Code		
SP02	Less Than Value = abs < Low Calibration Point	
Prepared By		
PPB		
Chemist		
PPB		
Analyst		
JIS		
Date Complete	The Sample Absorbance is < Detection Limit (the Low Calibration Poi	
11/18/98		
Analysis Date	v RESULT v	BLANK
10/30/98		
Analysis Time	Cr +6 µg/mL =	<0.258
12:50 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	11/18/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1034

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.189	0.187
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26742	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005343	Cr +6 µg/mL =	9.41E+02	9.31E+02
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002940	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*1		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
PPB			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
11/18/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
10/30/98			
Analysis Time	Cr +6 µg/mL =	9.41E+02	9.31E+02
12:50 PM			
Sample Point			
BY-105	Recovery % Difference =	1.08%	

Analyst:

JIS

Date: 11/18/98

Signature of Chemist:

PPB

Date:

SAMPLE.WB1 REV 1.

265101

1035

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.189
SPK	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.59
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
26742	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98005343	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T002940	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
PPB	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
11/18/98		
Analysis Date		
10/30/98		
Analysis Time	RESULT	
12:50 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	5.73E+00
BY-105	% SPIKE RECOVERY =	113.0%

Analyst:	JIS	Date: 11/18/98
Signature of Chemist:	PPB	Date:

SPIKE.WB1 REV 1.0

265101

1036

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.225	0.225
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26742	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005343	Cr +6 µg/mL =	1.12E+03	1.12E+03
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002965	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*1		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
PPB			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
11/18/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
10/30/98			
Analysis Time	Cr +6 µg/mL =	1.12E+03	1.12E+03
12:50 PM			
Sample Point			
BY-105	Recovery % Difference = 0.00%		

Analyst:	JIS	Date:	11/18/98
Signature of Chemist:		PPB	Date:
SAMPLE.WB1 REV 1.			
265101			

1037

WORKBOOK PAGE: SPIKE6

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.225
SPK	SAMPLE + SPIKE ABSORBANCE at 540 nm (S+S abs)	0.572
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
26742	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98005343	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T002965	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m]/(MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m]/(SSMF) (MSSV)	
PPB	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
11/18/98		
Analysis Date		
10/30/98		
Analysis Time	RESULT	
12:50 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	5.03E+00
BY-105	% SPIKE RECOVERY =	99.3%

Analyst:	JIS	Date: 11/18/98
Signature of Chemist:	PPB	Date:

SPIKE.WB1 REV 1.0

265101

1038

WORKBOOK PAGE: SAM7

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.220	0.218
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26742	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005343	Cr +6 µg/mL =	1.10E+03	1.09E+03
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002966	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*1		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
PPB			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
11/18/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
10/30/98			
Analysis Time	Cr +6 µg/mL =	1.10E+03	1.09E+03
12:50 PM			
Sample Point			
BY-105	Recovery % Difference =	0.92%	

Analyst:	JIS	Date:	11/18/98
Signature of Chemist:	<i>NA</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1039

LABCORE Completed Worklist Report for Worklist# 26743

Analyst: jis Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCV	0		CRVI-01	LIQUID	5.06	5.13	101.183 % Recovery	
2 CCB	0		CRVI-01	LIQUID	1	<0.258	Ratio	
3 SAMPLE	898T002967	0	CRVI-01	LIQUID	N/A	1.20e+3	1.00e-002 ug/mL	
4 DUP	898T002967	0	CRVI-01	LIQUID	1.20e+3	1.19e+3	0.837 RPD	
5 SPK	898T002967	0	CRVI-01	LIQUID	100	99.1	99.100 % Recovery	

Final page for worklist# 26743

Analyst Signature _____ Date _____

Cf. Quinn 12/2/98
Analyst Signature Date

B. Machelor 12/2/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26743

Analyst: JLS Instrument: SP02 Book# 49119-D

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002967 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T002967 0		CRVI-01	LIQUID		
5 SPK	S98T002967 0		CRVI-01	LIQUID		
6 SAMPLE	S98T002968 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: CRVI-01					
7 DUP	S98T002968 0		CRVI-01	LIQUID		
8 SPK	S98T002968 0		CRVI-01	LIQUID		
9 SAMPLE	S98T002969 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: CRVI-01					
10 DUP	S98T002969 0		CRVI-01	LIQUID		

Final page for worklist # 26743

Deleted
10/30/98

Analyst Signature JLS Date 10-27-98

Analyst Signature CL Date 10/2/98

Data Entry Comments:

S98T002968 & S98T002969 NOT Run Due
To HPT coverage!

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	CCV
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26743	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.13E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
STD		
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By		
CJO	Less Than Value = abs < Low Calibration Point	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
12/02/98		
Analysis Date	v RESULT v	CCV
10/27/98		
Analysis Time	Cr +6 µg/mL =	5.13E+00
02:20 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

STANDARD.WB1 REV 1.1

265101

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.001
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26743	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
BLANK
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/02/98
Analysis Date
10/27/98
Analysis Time
02:20 PM
Sample Point
BY-105

Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$

If the result is a < value (abs < LOW) :

Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$ RPD = $(|Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP| \cdot 100 /$
(AVERAGE OF THE SAM + DUP)

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1043

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.240	0.239
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26743	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005344	Cr +6 µg/mL =	1.20E+03	1.19E+03
Rerun	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
0	If the result is a < value (abs < LOW) :		
Sample Prep	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
N/A	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
Sample #	Less Than Value = abs < Low Calibration Point		
S98T002967			
Instrument Code			
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
12/02/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
10/27/98	Cr +6 µg/mL =	1.20E+03	1.19E+03
Analysis Time			
02:20 PM			
Sample Point			
BY-105	Recovery % Difference =	0.42%	

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1044

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.24
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.585
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
26743	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98005344	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T002967	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
CJO	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
12/02/98		
Analysis Date		
10/27/98		
Analysis Time	RESULT	
02:20 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	5.02E+00
BY-105	% SPIKE RECOVERY =	99.1%

Analyst:	JIS	Date: 12/02/98
Signature of Chemist:	PPB	Date:

SPIKE.WB1 REV 1.0

265101

LABCORE Completed Worklist Report for Worklist# 26844

Analyst: jis Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCV	0		CRVI-01	LIQUID	5.06	5.08	100.395 % Recovery	
2	CCB	0		CRVI-01	LIQUID	1	<0.258	Ratio	
3	SAMPLE	S98T003010	0	CRVI-01	LIQUID	N/A	1.11e+3	1.00e-002	ug/mL
4	DUP	S98T003010	0	CRVI-01	LIQUID	1.11e+3	1.11e+3	0.000	RPD
5	SAMPLE	S98T003011	0	CRVI-01	LIQUID	N/A	1.10e+3	1.00e-002	ug/mL
6	DUP	S98T003011	0	CRVI-01	LIQUID	1.10e+3	1.11e+3	0.905	RPD

Final page for worklist# 26844

Analyst Signature _____ Date _____

C. J. Owen 12/2/98
Analyst Signature Date

S. Machelon 12/2/98
Reviewer Signature Date

10/26/98 14:28
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26844 Analyst: JIS Instrument: SP02 Book# 49N19-DMethod: LA-265-101 Rev/Mod C.O

Worklist Comment: BY-105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T003010 0		CRVI-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T003010 0		CRVI-01	LIQUID		
5 SAMPLE	S98T003011 0		CRVI-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: CRVI-01					
6 DUP	S98T003011 0		CRVI-01	LIQUID		

Final page for worklist # 26844

Analyst Signature

Date

11-12-98

Analyst Signature

Date

12/2/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1047

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		CCV
Type	Sample Absorbance at 540 nm (abs)	0.381
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26844	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.08E+00
0		

Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)
N/A	If the result is a < value (abs < LOW) :
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)
CCV	

Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /
SP02	(AVERAGE OF THE SAM + DUP)

Prepared By	Less Than Value = abs < Low Calibration Point
CJO	

Chemist
PPB

Analyst
JIS

Date Complete
12/02/98

Analysis Date	v RESULT v
11/12/98	

Analysis Time	Cr +6 µg/mL =	CCV
02:30 PM		5.08E+00

Sample Point
BY-105

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:		PPB	Date:

STANDARD.WB1 REV 1.1

265101

1048

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26844	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
BLANK
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/02/98
Analysis Date
11/12/98
Analysis Time
02:30 PM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$\text{RPD} = (| \text{Cr +6 } \mu\text{g/mL SAM} - \text{Cr +6 } \mu\text{g/mL DUP} | \cdot 100 /$$

$$(\text{AVERAGE OF THE SAM} + \text{DUP}))$$

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	<i>N/A</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1049

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.223	0.223
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26844	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005448	Cr +6 µg/mL =	1.11E+03	1.11E+03
Rerun	$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] * (abs - a) (DDF)$		
0			
Sample Prep	If the result is a < value (abs < LOW) : $\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] * (LOW - a) (DDF)$		
N/A			
Sample #	$\text{RPD} = (\text{Cr +6 } \mu\text{g/mL SAM} - \text{Cr +6 } \mu\text{g/mL DUP} / \text{AVERAGE OF THE SAM + DUP}) * \text{Less Than Value} = abs < \text{Low Calibration Point}$		
S98T003010			
Instrument Code			
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
12/02/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
11/12/98			
Analysis Time	Cr +6 µg/mL =	1.11E+03	1.11E+03
02:30 PM			
Sample Point			
BY-105	Recovery % Difference =	0.00%	

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1050

WORKBOOK PAGE: SAM4

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.220	0.222
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26844	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005448	Cr +6 µg/mL =	1.10E+03	1.11E+03
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T003011	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
12/02/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
11/12/98			
Analysis Time	Cr +6 µg/mL =	1.10E+03	1.11E+03
02:30 PM			
Sample Point			
BY-105	Recovery % Difference = 0.92%		

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	<i>NA</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1051

LABCORE Completed Worklist Report for Worklist# 26845

Analyst: jis Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCV	0		CRVI-01	LIQUID	5.06	4.97	98.221 % Recovery	
2	CCB	0		CRVI-01	LIQUID	1	<0.258	Ratio	
3	SAMPLE	S98T003012	0	CRVI-01	LIQUID	N/A	1.09e+3	1.00e-002 ug/mL	
4	DUP	S98T003012	0	CRVI-01	LIQUID	1.09e+3	1.10e+3	0.913 RPD	
5	SPK	S98T003012	0	CRVI-01	LIQUID	100	95.3	95.300 % Recovery	
6	SAMPLE	S98T003013	0	CRVI-01	LIQUID	N/A	1.12e+3	1.00e-002 ug/mL	
7	DUP	S98T003013	0	CRVI-01	LIQUID	1.12e+3	1.13e+3	0.889 RPD	

Final page for worklist# 26845

Analyst Signature _____ Date _____

C. J. O'Brien 12/2/98
Analyst Signature Date

B. Brachet 12/2/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 26845

Analyst: UIS Instrument: SP02 Book# 49111-0

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 FOR CRVI-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCV				CRVI-01	LIQUID		
2	CCB				CRVI-01	LIQUID		
3	SAMPLE	S98T003012 0			CRVI-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CRVI-01						
4	DUP	S98T003012 0			CRVI-01	LIQUID		
5	SPK	S98T003012 0			CRVI-01	LIQUID		
6	SAMPLE	S98T003013 0			CRVI-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CRVI-01						
7	DUP	S98T003013 0			CRVI-01	LIQUID		

Final page for worklist # 26845

Jeff Salzman 11-11-98
Analyst Signature Date

C.J. Quinn 12/2/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		CCV
Type	Sample Absorbance at 540 nm (abs)	0.373
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26845	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	4.97E+00

0

Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)
N/A	If the result is a < value (abs < LOW) :
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)
CCV	

Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /
SP02	(AVERAGE OF THE SAM + DUP)

Prepared By	Less Than Value = abs < Low Calibration Point
CJO	

Chemist
PPB

Analyst
JIS

Date Complete
12/02/98

Analysis Date
11/11/98

Analysis Time
01:20 PM

Sample Point
BY-105

v RESULT v	CCV
Cr +6 µg/mL =	4.97E+00

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:		PPB	Date:

STANDARD.WB1 REV 1.1

265101

1054

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26845	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
Instrument Code
Prepared By
Chemist
Analyst
Date Complete
Analysis Date
Analysis Time
Sample Point

Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$

If the result is a < value (abs < LOW) :

Cr +6 µg/mL = $< [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$

$$RPD = (|Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP| \cdot 100 /$$

$$(AVERAGE OF THE SAM + DUP))$$

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:

Date:

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.218	0.220
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26845	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005449	Cr +6 µg/mL =	1.09E+03	1.10E+03

Rerun: 0
 Sample Prep: N/A
 Sample #: S98T003012
 Instrument Code: SP02
 Prepared By: CJO
 Chemist: PPB
 Analyst: JIS
 Date Complete: 12/02/98
 Analysis Date: 11/11/98
 Analysis Time: 01:20 PM
 Sample Point: BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|Cr +6 \mu\text{g/mL SAM} - Cr +6 \mu\text{g/mL DUP}| / \text{AVERAGE OF THE SAM} + \text{DUP}) \cdot 100$$

Less Than Value = abs < Low Calibration Point

v RESULT v

	SAMPLE	DUPLICATE
Cr +6 µg/mL =	1.09E+03	1.10E+03
Recovery % Difference =	0.92%	

Analyst:

Date:

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.218
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.551
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
26845	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98005449	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T003012	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
CJO	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
12/02/98		
Analysis Date		
11/11/98		
Analysis Time	RESULT	
01:20 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	4.83E+00
BY 105	% SPIKE RECOVERY =	95.3%

Analyst:	Date:
Signature of Chemist:	Date:

SPIKE.WB1 REV 1.0

265101

1057

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.225	0.227
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26845	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005449	Cr +6 µg/mL =	1.12E+03	1.13E+03
Rerun			
0	Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = $[(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$		
Sample #			
S98T003013	RPD = $(Cr +6 \mu g/mL SAM - Cr +6 \mu g/mL DUP / AVERAGE OF THE SAM + DUP) \cdot$		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
JIS			
Date Complete			
12/02/98			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
11/11/98			
Analysis Time	Cr +6 µg/mL =	1.12E+03	1.13E+03
01:20 PM			
Sample Point			
BY-105	Recovery % Difference =	0.90%	

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1058

LABCORE Completed Worklist Report for Worklist# 26984

Analyst: jjs Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY105 FOR CRVI-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCV	0		CRVI-01	LIQUID	5.06	4.97	98.221 %	Recovery
2 CCB	0		CRVI-01	LIQUID	1	<0.258		Ratio
3 SAMPLE	S98T002968	0	CRVI-01	LIQUID	N/A	1.01e+3	1.00e-002	ug/mL
4 DUP	S98T002968	0	CRVI-01	LIQUID	1.01e+3	9.86e+2		2.405 RPD
5 SPK	S98T002968	0	CRVI-01	LIQUID	100	95.0		95.000 % Recovery
6 SAMPLE	S98T002969	0	CRVI-01	LIQUID	N/A	1.24e+3	1.00e-002	ug/mL
7 DUP	S98T002969	0	CRVI-01	LIQUID	1.24e+3	1.26e+3		1.600 RPD

Final page for worklist# 26984

Analyst Signature _____ Date _____

C.J. Dineen 12/2/98
Analyst Signature Date

B. McChelton 12/2/98
Reviewer Signature Date

11/04/98 15:53
A-0004-1

HNF-1663 REV. 0

Page: 1

LABCORP Data Entry Template for Worklist# 26984

Analyst: JIS Instrument: SP02 Book# 49N19-D

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
3 SAMPLE	S98T002968 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: CRVI-01					
4 DUP	S98T002968 0		CRVI-01	LIQUID		
5 SPK	S98T002968 0		CRVI-01	LIQUID		
6 SAMPLE	S98T002969 0		CRVI-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested: CRVI-01					
7 DUP	S98T002969 0		CRVI-01	LIQUID		

Final page for worklist # 26984

Jay Solback 11-2-98
Analyst Signature Date

C. J. Pina 12/2/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1060

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	CCV
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
26984	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-02	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	4.97E+00

0

Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)
N/A	If the result is a < value (abs < LOW) :
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)

CCV	
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /
SP02	(AVERAGE OF THE SAM + DUP)

Prepared By	
CJO	Less Than Value = abs < Low Calibration Point

Chemist
PPB
Analyst
JIS

Date Complete
12/02/98

Analysis Date	v RESULT v	CCV
11/12/98		
Analysis Time	Cr +6 µg/mL =	4.97E+00
01:00 PM		
Sample Point		
BY-105		

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:		PPB	Date:

STANDARD.WB1 REV 1.1

265101

1061

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	BLANK
CCB	Sample Volume in mL (Vsamp)	1.00
Worklist	Method Factor (10 + Vsamp) (MF)	11
26984	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
CCB
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/02/98
Analysis Date
11/12/98
Analysis Time
01:00 PM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$\text{RPD} = (|\text{Cr +6 } \mu\text{g/mL SAM} - \text{Cr +6 } \mu\text{g/mL DUP}| \cdot 100 /$$

(AVERAGE OF THE SAM + DUP)

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL ≈	<0.258

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	<i>NA</i>	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1062

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.203	0.198
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26984	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005605	Cr +6 µg/mL =	1.01E+03	9.86E+02

Rerun
0
Sample Prep
N/A
Sample #
S98T002968
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/02/98
Analysis Date
11/12/98
Analysis Time
01:00 PM
Sample Point
BY-105

$$\text{Cr} + 6 \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (\text{abs} - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr} + 6 \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (\text{LOW} - a) (DDF)$$

$$\text{RPD} = (|\text{Cr} + 6 \mu\text{g/mL SAM} - \text{Cr} + 6 \mu\text{g/mL DUP}| / \text{AVERAGE OF THE SAM} + \text{DUP}) \cdot$$

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL =

	SAMPLE	DUPLICATE
	1.01E+03	9.86E+02

Recovery % Difference = 2.53%

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1063

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.203
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.536
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
26984	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98005605	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T002968	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
CJO	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
JIS		
Date Complete		
12/02/98		
Analysis Date		
11/12/98		
Analysis Time	RESULT	
01:00 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	4.81E+00
BY-105	% SPIKE RECOVERY =	95.0%

Analyst:	JIS	Date: 12/02/98
Signature of Chemist:	PPB	Date:

SPIKE.WB1 REV 1.0

265101

1064

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.248	0.253
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Work List	Method Factor (10 + Vsamp) (MF)	10.25	10.25
26984	Dilution Factor (DF)	101	101
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98005605	Cr +6 µg/mL =	1.24E+03	1.26E+03

Rerun
0
Sample Prep
N/A
Sample #
S98T002969
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
JIS
Date Complete
12/02/98
Analysis Date
11/12/98
Analysis Time
01:00 PM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|\text{Cr +6 } \mu\text{g/mL SAM} - \text{Cr +6 } \mu\text{g/mL DUP}| / \text{AVERAGE OF THE SAM + DUP}) \cdot 100$$

Less Than Value = abs < Low Calibration Point

v RESULT v

Cr +6 µg/mL =

	SAMPLE	DUPLICATE
	1.24E+03	1.26E+03
Recovery % Difference =	2.02%	

Analyst:	JIS	Date:	12/02/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1065

LABCORE Completed Worklist Report for Worklist# 27639

Analyst: pjm

Instrument: SP02

Book#: 49N19E

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 CRVI-01 cjo

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCV		0	CRVI-01	LIQUID	5.05	5.05	99.802 % Recovery	
2	CCB		0	CRVI-01	LIQUID	1	<.258	Ratio	
3	SAMPLE	S98T003385	0	CRVI-01	LIQUID	N/A	1.53e+3	1.00e-002 ug/mL	
4	DUP	S98T003385	0	CRVI-01	LIQUID	1.53e+3	1.55e+3	1.299 RPD	
5	SPK	S98T003385	0	CRVI-01	LIQUID	100	98.2	98.200 % Recovery	

Final page for worklist# 27639

Analyst Signature _____ Date _____

12/28/98
Analyst Signature Date

12/29/98
Reviewer Signature Date

12/18/98 08:35
A-0004-1

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 27639

Analyst: RM Instrument: SP02 _____ Book# 49N19-E

Method: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 CRVI-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	LIQUID		
2 CCB			CRVI-01	LIQUID		
~~3 BLNK-PREF~~ ^{RM} <u>12/19/98</u>			CRVI-01	LIQUID		
4 SAMPLE	S98T003385 0		CRVI-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CRVI-01						
5 DUP	S98T003385 0		CRVI-01	LIQUID		
6 SPK	S98T003385 0		CRVI-01	LIQUID		

Final page for worklist # 27639

RM 12/19/98
Analyst Signature Date

C.D. O'Brien 12/28/98
Analyst Signature Date

Data Entry Comments: Direct sample. No prep blank run.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1067

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE****CCV**

Type	Sample Absorbance at 540 nm (abs)	0.379
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27639	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.05E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By	Less Than Value = abs < Low Calibration Point	
CJO		
Chemist		
PPB		
Analyst		
PJM		
Date Complete		
12/28/98		
Analysis Date	v RESULT v	CCV
12/19/98		
Analysis Time	Cr +6 µg/mL =	5.05E+00
08:00 PM		
Sample Point		
BY-105		

Analyst:

PJM

Date:

12/28/98

Signature of Chemist:

NA

PPB

Date:

STANDARD.WB1 REV 1.1

265101

1068

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27639	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
CCB
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
PJM
Date Complete
12/28/98
Analysis Date
12/19/98
Analysis Time
08:00 PM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|Cr +6 \mu\text{g/mL SAM} - Cr +6 \mu\text{g/mL DUP}| \cdot 100 /$$

(AVERAGE OF THE SAM + DUP)

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:	PJM	Date:	12/28/98
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)

LIQUID/SOLID SAMPLE

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.306	0.309
SAMPLE	Sample Volume in mL (Vsamp)	0.25	0.25
Worklist	Method Factor (10 + Vsamp) (MF)	10.25	10.25
27639	Dilution Factor (DF)	101	101
Test Codes	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006266	Cr +6 µg/mL =	1.53E+03	1.55E+03
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T003385	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
PJM			
Date Complete			
12/28/98			
Analysis Date	v RESULT v		
12/19/98			
Analysis Time	Cr +6 µg/mL =		
08:00 PM			
Sample Point			
BY-105	Recovery % Difference = 0.98%		

Analyst:	NA	PJM	Date: 12/28/98
Signature of Chemist:	NA	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1070

WORKBOOK PAGE: SPIKE4

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Sample Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.306
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.642
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	0.250
27639	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	0.250
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	10.25
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	11.25
LIQUID	Sample Dilution Factor (DF)	101
Batch Number	Digest Dilution Factor (DDF)	1
98006266	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (ug/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
N/A	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T003385	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
CJO	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
PJM		
Date Complete		
12/28/98		
Analysis Date		
12/19/98		
Analysis Time	RESULT	
08:00 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	4.98E+00
BY-105	% SPIKE RECOVERY =	98.2%

Analyst:	PJM	Date: 12/28/98
Signature of Chemist:	PPB	Date:

SPIKE.WB1 REV 1.0

265101

1071

LABCORE Completed Worklist Report for Worklist# 27689

Analyst: slh Instrument: SP02 Book#: _____

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY105 FOR CRVI-01 RTS

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCV	0		CRVI-01	SOLID	5.06	5.10	100.791 % Recovery	
2	CCB	0		CRVI-01	SOLID	1	<258.16	Ratio	
3	BLNK-PREP	0		CRVI-01	SOLID	1	<27.96	ug/g	
4	SAMPLE	S98T003390	0 W	CRVI-01	SOLID	N/A	1.25e2	1.00e-002 ug/g	
5	DUP	S98T003390	0 W	CRVI-01	SOLID	1.25e2	1.26e2	0.797 RPD	
6	SPK	S98T003390	0 W	CRVI-01	SOLID	100	100.9	100.900 % Recovery	
7	SAMPLE	S98T003395	0 W	CRVI-01	SOLID	N/A	1.03e2	1.00e-002 ug/g	
8	DUP	S98T003395	0 W	CRVI-01	SOLID	1.03e2	1.01e2	1.961 RPD	
9	SPK	S98T003395	0 W	CRVI-01	SOLID	100	100.5	100.500 % Recovery	
10	SAMPLE	S98T003400	0 W	CRVI-01	SOLID	N/A	8.47e1	1.00e-002 ug/g	
11	DUP	S98T003400	0 W	CRVI-01	SOLID	8.47e1	6.80e1	21.873 RPD	
12	SPK	S98T003400	0 W	CRVI-01	SOLID	100	95.3	95.300 % Recovery	

Final page for worklist# 27689

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 1/7/99

LABCORE Data Entry Template for Worklist# 27689

Analyst: B.H. Instrument: SP02 _____ Book# 49N19-EMethod: LA-265-101 Rev/Mod C-0

Worklist Comment: BY105 FOR CRVI-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCV			CRVI-01	SOLID		
2 CCB			CRVI-01	SOLID		
3 BLNK-PREP			CRVI-01	SOLID		
4 SAMPLE	S98T003390 0 W		CRVI-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: CRVI-01					
5 DUP	S98T003390 0 W		CRVI-01	SOLID		
6 SPK	S98T003390 0 W		CRVI-01	SOLID		
7 SAMPLE	S98T003395 0 W		CRVI-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: CRVI-01					
8 DUP	S98T003395 0 W		CRVI-01	SOLID		
9 SPK	S98T003395 0 W		CRVI-01	SOLID		
10 SAMPLE	S98T003400 0 W		CRVI-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: CRVI-01					
11 DUP	S98T003400 0 W		CRVI-01	SOLID		
12 SPK	S98T003400 0 W		CRVI-01	SOLID		

Final page for worklist # 27689

Analyst Signature

Date

12-30-98

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	CCV
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27689	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
SOLID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.10E+00
0		

Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)
N/A	If the result is a < value (abs < LOW) :
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)
CCV	

Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /
SP02	(AVERAGE OF THE SAM + DUP)

Prepared By	Less Than Value = abs < Low Calibration Point
MF	

Chemist
PPB
Analyst
SLH

Date Complete
01/04/99

Analysis Date	v RESULT v	CCV
12/30/98		
Analysis Time	Cr +6 µg/mL =	5.10E+00
05:00 PM		
Sample Point		
BY105		

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	PPB	Date:	

STANDARD.WB1 REV 1.1

265101

1074

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27689	Dilution Factor (DF)	1
Test Code	Digest g/L (DGL)	1
CRVI-01	Low Calibration abs point (LOW)	0.022
Matrix	High Calibration abs point	0.831
SOLID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/g =	<258.161
Rerun		
0	Cr +6 µg/g = $[(MF*DF) / (b*VSAMP)] * (abs - a) (1000 / DGL)$	
Sample Prep	If the result is a < value (abs < LOW) :	
N/A	Cr +6 µg/g = $< [(MF*DF) / (b*VSAMP)] * (LOW - a) (1000 / DGL)$	
Sample #	RPD = $ Cr +6 \mu g/g \text{ SAM} - Cr +6 \mu g/g \text{ DUP} * 100 /$	
CCB	(AVERAGE OF THE SAM + DUP)	
Instrument Code		
SPO2	Less Than Value = Sample Absorbance < Low Calibration Point	
Prepared By		
MF		
Chemist		
PPB		
Analyst		
SLH		
Date Complete	The Sample Absorbance is < Detection Limit (the Low Calibration Poi	
01/04/99		
Analysis Date	v RESULT v	BLANK
12/30/98		
Analysis Time	Cr +6 µg/g =	<258.161
05:00 PM		
Sample Point		
BY105		

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1075

WORKBOOK PAGE: BLANK3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
BLNK-PREP	Sample Volume in mL (Vsamp)	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12
27689	Dilution Factor (DF)	1
Test Code	Digest g/L (DGL)	5.036
CRVI-01	Low Calibration abs point (LOW)	0.022
Matrix	High Calibration abs point	0.831
SOLID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
98006329	Cr +6 µg/g =	<27.962

Rerun	
0	$Cr +6 \mu g/g = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (1000 / DGL)$
Sample Prep	If the result is a < value (abs < LOW) :
N/A	$Cr +6 \mu g/g = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (1000 / DGL)$
Sample #	$RPD = Cr +6 \mu g/g SAM - Cr +6 \mu g/g DUP \cdot 100 /$
BLNK-PREP	(AVERAGE OF THE SAM + DUP)
Instrument Code	
SP02	Less Than Value = Sample Absorbance < Low Calibration Point

Prepared By	
MF	
Chemist	
MJL	
Analyst	
SLH	
Date Complete	
01/04/99	
Analysis Date	
12/30/98	
Analysis Time	
05:00 PM	
Sample Point	
BY105	

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v

Cr +6 µg/g =

BLANK

<27.962

Analyst: SLH Date: 01/04/99

Signature of Chemist: MJL Date:

SAMPLE.WB1 REV 1.

265101

1076

WORKBOOK PAGE: SAM4

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.089	0.091
SAMPLE	Sample Volume in mL (Vsamp)	2.00	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12	12
27689	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	5.036	5.133
CRVI-01	Low Calibration abs point (LOW)	0.022	0.022
Matrix	High Calibration abs point	0.831	0.831
SOLID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006329	Cr +6 µg/g =	1.25E+02	1.26E+02

Rerun			
0	Cr +6 µg/g = [(MF*DF) / (b*VSAMP)] * (abs - a) (1000 / DGL)		
Sample Prep	If the result is a < value (abs < LOW) :		
H2ODIG01	Cr +6 µg/g = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (1000 / DGL)		
Sample #			
S98T003390	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = Sample Absorbance < Low Calibration Point		
SP02			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
01/04/99			
Analysis Date	v RESULT v		
12/30/98			
Analysis Time	Cr +6 µg/g =	SAMPLE	DUPLICATE
05:00 PM		1.25E+02	1.26E+02
Sample Point			
BY105	Recovery % Difference =	0.38%	

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	MJL	Date:	

SAMPLE.WB1 REV 1.

265101

1077

WORKBOOK PAGE: SPIKE5

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.089
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.405
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	2.000
27689	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	2.000
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	12
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	13
SOLID	Sample Dilution Factor (DF)	1
Batch Number	Digest g/L (DGL)	5.036
98006329	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
H2ODIG01	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T003390	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
MF	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
01/04/99		
Analysis Date		
12/30/98		
Analysis Time	RESULT	
05:00 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	5.12E+00
BY105	% SPIKE RECOVERY =	100.9%

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	MJL	Date:	

SPIKE.WB1 REV 1.0

265101

1078

WORKBOOK PAGE: SAM6

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.072	0.074
SAMPLE	Sample Volume in mL (Vsamp)	2.00	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12	12
27689	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	4.942	5.158
CRVI-01	Low Calibration abs point (LOW)	0.022	0.022
Matrix	High Calibration abs point	0.831	0.831
SOLID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006329	Cr +6 µg/g =	1.03E+02	1.01E+02

Re-run
0
Sample Prep
H2ODIG01
Sample #
S98T003395
Instrument Code
SP02
Prepared By
MF
Chemist
MJL
Analyst
SLH
Date Complete
01/04/99
Analysis Date
12/30/98
Analysis Time
05:00 PM
Sample Point
BY105

$$\text{Cr +6 } \mu\text{g/g} = [(MF \cdot DF) / (b \cdot VSAMP)] * (abs - a) (1000 / DGL)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/g} = < [(MF \cdot DF) / (b \cdot VSAMP)] * (LOW - a) (1000 / DGL)$$

$$RPD = |\text{Cr +6 } \mu\text{g/g SAM} - \text{Cr +6 } \mu\text{g/g DUP}| / \text{AVERAGE OF THE SAM + DUP} * 100$$

Less Than Value = Sample Absorbance < Low Calibration Point

v RESULT v

	SAMPLE	DUPLICATE
Cr +6 µg/g =	1.03E+02	1.01E+02
Recovery % Difference =	1.43%	

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	MJL	Date:	

SAMPLE.WB1 REV 1.

265101

1079

WORKBOOK PAGE: SPIKE7

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.072
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.388
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	2.000
27689	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	2.000
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	12
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	13
SOLID	Sample Dilution Factor (DF)	1
Batch Number	Digest g/L (DGL)	4.942
98006329	SPIKE DILUTION FACTOR (SDF)	1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.07
0		
Sample Prep	LOW CALIBRATION POINT	0.022
H2ODIG01	HIGH CALIBRATION POINT	0.831
Sample #	Y-INTERCEPT (b)	0.00276
S98T003395	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
MF	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
MJL		
Analyst		
SLH		
Date Complete		
01/04/99		
Analysis Date		
12/30/98		
Analysis Time	RESULT	
05:00 PM	QC ACTUAL =	5.07E+00
Sample Point	QC FOUND =	5.10E+00
BY105	% SPIKE RECOVERY =	100.5%

Analyst:	SLH	Date: 01/04/99
Signature of Chemist:	MJL	Date:

SPIKE.WB1 REV 1.0

265101

1080

WORKBOOK PAGE: SAM8

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.084	0.068
SAMPLE	Sample Volume in mL (Vsamp)	3.00	3.00
Work List	Method Factor (10 + Vsamp) (MF)	13	13
27689	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	5.068	5.072
CRVI-01	Low Calibration abs point (LOW)	0.022	0.022
Matrix	High Calibration abs point	0.831	0.831
SOLID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006329	Cr +6 µg/g =	8.47E+01	6.80E+01
Rerun			
0	Cr +6 µg/g = [(MF*DF) / (b*VSAMP)] * (abs - a) (1000 / DGL)		
Sample Prep	If the result is a < value (abs < LOW) :		
H2ODIG01	Cr +6 µg/g = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (1000 / DGL)		
Sample #	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
S98T003400	Less Than Value = Sample Absorbance < Low Calibration Point		
Instrument Code			
SP02			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
01/04/99			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
12/30/98			
Analysis Time	Cr +6 µg/g =	8.47E+01	6.80E+01
05:00 PM			
Sample Point			
BY105	Recovery % Difference = 21.92%		

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	MJL	Date:	

SAMPLE.WB1 REV 1.

265101

1081

WORKBOOK PAGE: SPIKE9

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

	SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs) 0.084
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs) 0.361
Work List	SAMPLE VOLUME (Sample) in mL (Vsam) 3.000
27689	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam) 3.000
Test Code	VOLUME of SPIKE in mL (VSPK) 1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF) 13
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF) 14
SOLID	Sample Dilution Factor (DF) 1
Batch Number	Digest g/L (DGL) 5.068
98006329	SPIKE DILUTION FACTOR (SDF) 1
Rerun	SPIKE STANDARD CONCENTRATION (µg/mL) 5.07
0	
Sample Prep	LOW CALIBRATION POINT 0.022
H2ODIG01	HIGH CALIBRATION POINT 0.831
Sample	Y-INTERCEPT (b) 0.00276
S98T003400	SLOPE (m) 0.8198
Instrument Code	
SPO2	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)
MF	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100
Chemist	
MJL	
Analyst	
SLH	
Date Complete	
01/04/99	
Analysis Date	
12/30/98	
Analysis Time	RESULT
05:00 PM	QC ACTUAL = 5.07E+00
Sample Point	QC FOUND = 4.83E+00
BY105	% SPIKE RECOVERY = 95.3%

Analyst:	SLH	Date:	01/04/99
Signature of Chemist:	MJL	Date:	

SPIKE.WB1 REV 1.0

265101

1082

12/30/98 14:55
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27737

Analyst: RJM Instrument: SP02 Book# 49219-FMethod: LA-265-101 Rev/Mod C-0

Worklist Comment: BY-105 FOR CRVI-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCV				CRVI-01	SOLID		
2	CCB				CRVI-01	SOLID		
3	BLNK-PREP				CRVI-01	SOLID		
4	SAMPLE	S98T002934 0 W			CRVI-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: CRVI-01						
5	DUP	S98T002934 0 W			CRVI-01	SOLID		
6	SAMPLE	S98T002935 0 W			CRVI-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: CRVI-01						
7	DUP	S98T002935 0 W			CRVI-01	SOLID		
8	SPK	S98T002935 0 W			CRVI-01	SOLID		

Final page for worklist # 27737

RJM
 Analyst Signature 1/5/99
 Date

Mary Trang
 Analyst Signature 1-8-99
 Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Completed Worklist Report for Worklist# 27737

Analyst: pjm

Instrument: SP02

Book#: 49N19F

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY-105 FOR CRVI-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCV	0	CRVI-01	SOLID	5.06	5.10	100.791 % Recovery	
2 CCB	0	CRVI-01	SOLID	1	<258.161	Ratio	
3 BLNK-PREP	0	CRVI-01	SOLID	1	<258.161	Ratio	
4 SAMPLE	S98T002934 0 W	CRVI-01	SOLID	N/A	2.25e2	1.00e-002 ug/g	
5 DUP	S98T002934 0 W	CRVI-01	SOLID	2.25e2	2.08e2	7.852 RPD	
6 SAMPLE	S98T002935 0 W	CRVI-01	SOLID	N/A	1.76e2	1.00e-002 ug/g	
7 DUP	S98T002935 0 W	CRVI-01	SOLID	1.76e2	2.03e2	14.248 RPD	
8 SPK	S98T002935 0 W	CRVI-01	SOLID	100	96.6	96.600 % Recovery	

Final page for worklist# 27737

Analyst Signature _____ Date _____

Mary Inoué 1-8-99
Analyst Signature _____ Date _____

B. Michael 1/18/99
Reviewer Signature _____ Date _____

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		CCV
Type	Sample Absorbance at 540 nm (abs)	0.383
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27737	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
SOLID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.10E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = [(Cr +6 µg/mL SAM - Cr +6 µg/mL DUP] *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By		
MF	Less Than Value = abs < Low Calibration Point	
Chemist		
PPB		
Analyst		
PJM		
Date Complete		
01/08/99		
Analysis Date	v RESULT v	CCV
01/05/99		
Analysis Time	Cr +6 µg/mL =	5.10E+00
03:30 PM		
Sample Point		
BY-105		

Analyst:	<i>NA</i>	PJM	Date: 01/08/99
Signature of Chemist:	<i>NA</i>	PPB	Date:

STANDARD.WB1 REV 1.1

265101

1085

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		BLANK
Type	Sample Absorbance at 540 nm (abs)	0.000
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
27737	Dilution Factor (DF)	1
Test Code	Digest g/L (DGL)	1
CRVI-01	Low Calibration abs point (LOW)	0.022
Matrix	High Calibration abs point	0.831
SOLID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/g =	<258.161

Rerun
0
Sample Prep
N/A
Sample #
CCB
Instrument Code
SP02
Prepared By
MF
Chemist
PPB
Analyst
PJM
Date Complete
01/08/99
Analysis Date
01/05/99
Analysis Time
03:30 PM
Sample Point
BY-105

$$\text{Cr} + 6 \mu\text{g/g} = [(MF \cdot DF) / (b \cdot VSAMP)] * (\text{abs} - a) (1000 / \text{DGL})$$

If the result is a < value (abs < LOW) :

$$\text{Cr} + 6 \mu\text{g/g} = < [(MF \cdot DF) / (b \cdot VSAMP)] * (\text{LOW} - a) (1000 / \text{DGL})$$

$$\text{RPD} = |\text{Cr} + 6 \mu\text{g/g SAM} - \text{Cr} + 6 \mu\text{g/g DUP}| * 100 /$$

(AVERAGE OF THE SAM + DUP)

Less Than Value = Sample Absorbance < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/g =	<258.161

Analyst:	<i>NA</i>	PJM	Date: 01/08/99
Signature of Chemist:		PPB	Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: BLANK3

Cr+6 : LA-265-101 (C-0)

LIQUID/SOLID SAMPLE

Type	Sample Absorbance at 540 nm (abs)	BLANK
BLNK-PREP	Sample Volume in mL (Vsamp)	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12
27737	Dilution Factor (DF)	1
Test Code	Digest g/L (DGL)	5.118
CRVI-01	Low Calibration abs point (LOW)	0.022
Matrix	High Calibration abs point	0.831
SOLID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
98006373	Cr +6 µg/g =	<27.514

Rerun
0
Sample Prep
N/A
Sample #
BLNK-PREP
Instrument Code
SP02
Prepared By
MF
Chemist
PPB
Analyst
PJM
Date Complete
01/08/99
Analysis Date
01/05/99
Analysis Time
03:30 AM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/g} = [(MF \cdot DF) / (b \cdot VSAMP)] * (abs - a) (1000 / DGL)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/g} = < [(MF \cdot DF) / (b \cdot VSAMP)] * (LOW - a) (1000 / DGL)$$

$$\text{RPD} = |\text{Cr +6 } \mu\text{g/g SAM} - \text{Cr +6 } \mu\text{g/g DUP}| * 100 /$$

(AVERAGE OF THE SAM + DUP)

Less Than Value = Sample Absorbance < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/g =	<27.514

Analyst:	NA	PJM	Date: 01/08/99
Signature of Chemist:	NA	PPB	Date:

SAMPLE.WB1 REV 1.

265101

1087

WORKBOOK PAGE: SAM4

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.160	0.147
SAMPLE	Sample Volume in mL (Vsamp)	2.00	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12	12
27737	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	5.118	5.086
CRVI-01	Low Calibration abs point (LOW)	0.022	0.022
Matrix	High Calibration abs point	0.831	0.831
SOLID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006373	Cr +6 µg/g =	2.25E+02	2.08E+02
Rerun	$\text{Cr +6 } \mu\text{g/g} = \frac{[(\text{MF} \cdot \text{DF}) / (\text{b} \cdot \text{VSAMP})] \cdot (\text{abs} - \text{a}) \cdot (1000 / \text{DGL})}{\text{If the result is a} < \text{value (abs} < \text{LOW)} :}$ $\text{Cr +6 } \mu\text{g/g} = \frac{[(\text{MF} \cdot \text{DF}) / (\text{b} \cdot \text{VSAMP})] \cdot (\text{LOW} - \text{a}) \cdot (1000 / \text{DGL})}{\text{RPD} = \text{Cr +6 } \mu\text{g/g SAM} - \text{Cr +6 } \mu\text{g/g DUP} / \text{AVERAGE OF THE SAM} + \text{DUP} \cdot 100}$ Less Than Value = Sample Absorbance < Low Calibration Point		
0			
Sample Prep			
H2ODIG01			
Sample #			
S98T002934			
Instrument Code			
SP02			
Prepared By			
MF			
Chemist			
PPB			
Analyst			
PJM			
Date Complete			
01/08/99			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
01/05/99	Cr +6 µg/g =	2.25E+02	2.08E+02
Analysis Time			
03:30 PM			
Sample Point			
BY-105	Recovery % Difference =	8.00%	

Analyst:	<i>NA</i>	PJM	Date:	01/08/99
Signature of Chemist:		PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1088

WORKBOOK PAGE: SAM5

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.122	0.141
SAMPLE	Sample Volume in mL (Vsamp)	2.00	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12	12
27737	Dilution Factor (DF)	1	1
Test Code	Digest g/L (DGL)	4.96	4.974
CRVI-01	Low Calibration abs point (LOW)	0.022	0.022
Matrix	High Calibration abs point	0.831	0.831
SOLID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
98006373	Cr +6 µg/g =	1.76E+02	2.03E+02
Rerun			
0	Cr +6 µg/g = [(MF*DF) / (b*VSAMP)] * (abs - a) (1000 / DGL)		
Sample Prep	If the result is a < value (abs < LOW) :		
H2ODIG01	Cr +6 µg/g = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (1000 / DGL)		
Sample #			
S98T002935	RPD = Cr +6 µg/g SAM - Cr +6 µg/g DUP / AVERAGE OF THE SAM + DUP * 100		
Instrument Code	Less Than Value = Sample Absorbance < Low Calibration Point		
SP02			
Prepared By			
MF			
Chemist			
PPB			
Analyst			
PJM			
Date Complete			
01/08/99			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
01/05/99	Cr +6 µg/g =	1.76E+02	2.03E+02
Analysis time			
03:30 PM			
Sample Point			
BY-105	Recovery % Difference =	14.48%	

Analyst:

NA

PJM

Date: 01/08/99

Signature of Chemist:

PPB

Date:

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SPIKE6

Cr+6 : LA-265-101 (C-0)

HEXAVALENT CHROMIUM - SPIKE

		SPIKE
Type	SAMPLE ABSORBANCE at 540 nm (SAM abs)	0.122
SPK	SAMPLE + SPIKE ABSORBANCE at 540 n (S+S abs)	0.421
Work List	SAMPLE VOLUME (Sample) in mL (Vsam)	2.000
27737	SAMPLE VOLUME (Spiked Sample) in mL (S+S Vsam)	2.000
Test Code	VOLUME of SPIKE in mL (VSPK)	1.000
CRVI-01	Total Vol Sample Vial (10 + Vsam) (MF)	12
Matrix	Total Vol Sam+Spk Vial (10+Vsam+VSPK) (SSMF)	13
SOLID	Sample Dilution Factor (DF)	1
Batch Number	Digest g/L (DGL)	4.96
98006373	SPIKE DILUTION FACTOR (SDF)	1
Retun	SPIKE STANDARD CONCENTRATION (µg/mL)	5.06
0		
Sample Prep	LOW CALIBRATION POINT	0.022
H2ODIG01	HIGH CALIBRATION POINT	0.831
Sample	Y-INTERCEPT (b)	0.00276
S98T002935	SLOPE (m)	0.8198
Instrument Code		
SP02	ug Cr +6 in SAMP VIAL = [(SAM abs - b)/m](MF) (MSV)	
Prepared By	ug Cr +6 in S+S VIAL = [(S+S abs - b) / m](SSMF) (MSSV)	
MF	% Spike Rec. = [(MSSV - MSV) / (SPIKE VALUE * VSPK / SDF)]*100	
Chemist		
PPB		
Analyst		
PJM		
Date Complete		
01/08/99		
Analysis Date		
01/05/99		
Analysis Time	RESULT	
03:30 PM	QC ACTUAL =	5.06E+00
Sample Point	QC FOUND =	4.89E+00
BY-105	% SPIKE RECOVERY =	96.6%

Analyst:	NA	PJM	Date: 01/08/99
Signature of Chemist:		PPB	Date:

SPIKE.WB1 REV 1.0

265101

1090

LABCORE Completed Worklist Report for Worklist# 28075

Analyst: pjm

Instrument: SP02

Book#: 49N19F

Method: LA-265-101 Rev/Mod _____

Worklist Comment: BY105 FB FOR CRVI-06 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCV	0	CRVI-01	LIQUID	5.06	5.10	100.791 % Recovery	
2 CCB	0	CRVI-01	LIQUID	1	<0.258	Ratio	
3 SAMPLE	998T002391 0	CRVI-01	LIQUID	N/A	<0.141	1.00e-002 ug/mL	
4 DUP	998T002391 0	CRVI-01	LIQUID	<0.141	<0.141	RPD	

Final page for worklist# 28075

Analyst Signature _____ Date _____

2/1/99
Analyst Signature _____ Date _____

2/1/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28075

Analyst: RM Instrument: SP02 _____ Book# 49N19F

Method: LA-265-101 Rev/Mod C-1

Worklist Comment: BY105 FB FOR CRVI-06 RLC

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCV				CRVI-01	LIQUID		
2	CCB				CRVI-01	LIQUID		
3	SAMPLE	S98T002391 0			CRVI-01	LIQUID	98000436	BY-105 (2)
					Analytes Requested: CRVI-01			
4	DUP	S98T002391 0			CRVI-01	LIQUID		

Final page for worklist # 28075

RM 1/28/99
Analyst Signature Date

C. J. Quinn 2/1/99
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	CCV
CCV	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
28075	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Standard Value	5.06E+00
Matrix	Low Calibration Point (LOW)	0.022
LIQUID	High Calibration Point"	0.831
Batch Number	Y-Intercept (a)	0.00276
	Slope (b)	0.8198
Rerun	Cr +6 µg/mL =	5.10E+00
0		
Sample Prep	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)	
N/A	If the result is a < value (abs < LOW) :	
Sample #	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)	
CCV		
Instrument Code	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP *100) /	
SP02	(AVERAGE OF THE SAM + DUP)	
Prepared By	Less Than Value = abs < Low Calibration Point	
CJO		
Chemist		
PPB		
Analyst		
PJM		
Date Complete		
02/01/99		
Analysis Date	v RESULT v	CCV
01/28/99		
Analysis Time	Cr +6 µg/mL =	5.10E+00
02:40 AM		
Sample Point		
BY-105		

Analyst:	<i>NA</i>	PJM	Date: 02/01/99
Signature of Chemist:		PPB	Date:

STANDARD.WB1 REV 1.1

265101

WORKBOOK PAGE: BLANK2

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

Type	Sample Absorbance at 540 nm (abs)	BLANK
CCB	Sample Volume in mL (Vsamp)	1.00
Work List	Method Factor (10 + Vsamp) (MF)	11
28075	Dilution Factor (DF)	1
Test Code	Digest Dilution Factor (DDF)	1
CRVI-01	Low Calibration Point (LOW)	0.022
Matrix	High Calibration Point	0.831
LIQUID	Y-Intercept (a)	0.00276
Batch Number	Slope (b)	0.8198
	Cr +6 µg/mL =	<0.258

Rerun
0
Sample Prep
N/A
Sample #
BLANK
Instrument Code
SP02
Prepared By
CJO
Chemist
PPB
Analyst
PJM
Date Complete
02/01/99
Analysis Date
01/28/99
Analysis Time
02:40 AM
Sample Point
BY-105

$$\text{Cr +6 } \mu\text{g/mL} = [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (abs - a) (DDF)$$

If the result is a < value (abs < LOW) :

$$\text{Cr +6 } \mu\text{g/mL} = < [(MF \cdot DF) / (b \cdot VSAMP)] \cdot (LOW - a) (DDF)$$

$$RPD = (|Cr +6 \mu\text{g/mL SAM} - Cr +6 \mu\text{g/mL DUP}| \cdot 100 / \text{AVERAGE OF THE SAM + DUP})$$

Less Than Value = abs < Low Calibration Point

The Sample Absorbance is < Detection Limit (the Low Calibration Poi

v RESULT v	BLANK
Cr +6 µg/mL =	<0.258

Analyst:	PJM	Date:	02/01/99
Signature of Chemist:	PPB	Date:	

SAMPLE.WB1 REV 1.

265101

WORKBOOK PAGE: SAM3

Cr+6 : LA-265-101 (C-0)**LIQUID/SOLID SAMPLE**

		SAMPLE	DUPLICATE
Type	Sample Absorbance at 540 nm (abs)	0.001	0.000
SAMPLE	Sample Volume in mL (Vsamp)	2.00	2.00
Work List	Method Factor (10 + Vsamp) (MF)	12	12
28075	Dilution Factor (DF)	1	1
Test Code	Digest Dilution Factor (DDF)	1	1
CRVI-01	Low Calibration Point (LOW)	0.022	0.022
Matrix	High Calibration Point	0.831	0.831
LIQUID	Y-Intercept (a)	0.00276	0.00276
Batch Number	Slope (b)	0.8198	0.8198
99000347	Cr +6 µg/mL =	<0.141	<0.141
Rerun			
0	Cr +6 µg/mL = [(MF*DF) / (b*VSAMP)] * (abs - a) (DDF)		
Sample Prep	If the result is a < value (abs < LOW) :		
N/A	Cr +6 µg/mL = < [(MF*DF) / (b*VSAMP)] * (LOW - a) (DDF)		
Sample #			
S98T002391	RPD = (Cr +6 µg/mL SAM - Cr +6 µg/mL DUP / AVERAGE OF THE SAM + DUP)*		
Instrument Code	Less Than Value = abs < Low Calibration Point		
SP02			
Prepared By			
CJO			
Chemist			
PPB			
Analyst			
PJM			
Date Complete	Both Absorbances are < Detection Limit (the Low Calibration Point)		
02/01/99			
Analysis Date	v RESULT v	SAMPLE	DUPLICATE
01/28/99			
Analysis Time	Cr +6 µg/mL =	<0.141	<0.141
02:40 AM			
Sample Point			
BY-105			

Analyst:	<i>NA</i>	PJM	Date:	02/01/99
Signature of Chemist:		PPB	Date:	

SAMPLE.WB1 REV 1.

265101

1095

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 4 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 1096 THRU 1449)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	ION CHROMATOGRAPHIC ANALYSIS, IC WORKLIST 25891 THRU 27692		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 7
of 17
Cont.

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

HNF-1663 REV.0

LABCORE Completed Worklist Report for Worklist# 25891

Analyst: adp

Instrument: IC40S1

Book#: 20N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	01C-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	01C-QC CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	01C-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	01C-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	01C-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	01C-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	01C-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	01C-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 ICV	0	01C-QC F	QC	5.30e1	5.88e+01	99.661 % Recovery	
2 ICV	0	01C-QC CL	QC	8.00e1	7.21e+01	90.125 % Recovery	
2 ICV	0	01C-QC NO2	QC	5.35e2	5.12e+02	95.701 % Recovery	
2 ICV	0	01C-QC BR	QC	5.86e2	5.38e+02	91.809 % Recovery	
2 ICV	0	01C-QC NO3	QC	5.92e2	5.30e+02	89.527 % Recovery	
2 ICV	0	01C-QC P04	QC	5.47e2	4.76e+02	87.020 % Recovery	
2 ICV	0	01C-QC S04	QC	6.38e2	5.88e+02	92.163 % Recovery	
2 ICV	0	01C-QC OXALATE2	QC	5.57e2	5.32e+02	95.512 % Recovery	
3 CCV	0	01C-QC F	QC	6.40e1	6.07e+01	94.844 % Recovery	
3 CCV	0	01C-QC CL	QC	9.00e1	8.80e+01	97.778 % Recovery	
3 CCV	0	01C-QC NO2	QC	5.49e2	5.18e+02	94.353 % Recovery	
3 CCV	0	01C-QC BR	QC	6.30e2	6.03e+02	95.714 % Recovery	
3 CCV	0	01C-QC NO3	QC	6.98e2	6.82e+02	97.708 % Recovery	
3 CCV	0	01C-QC P04	QC	6.32e2	5.79e+02	91.614 % Recovery	
3 CCV	0	01C-QC S04	QC	6.99e2	6.68e+02	95.565 % Recovery	
3 CCV	0	01C-QC OXALATE2	QC	5.40e2	5.19e+02	96.111 % Recovery	
4 SAMPLE	S98T002391	0 01C-01 F-02	LIQUID	N/A <	1.200e-02	1.20e-002	ug/mL
4 SAMPLE	S98T002391	0 01C-01 CL-02	LIQUID	N/A <	3.300e-02	1.70e-002	ug/mL
4 SAMPLE	S98T002391	0 01C-01 NO2-02	LIQUID	N/A <	1.080e-01	0.108	ug/mL
4 SAMPLE	S98T002391	0 01C-01 BR-02	LIQUID	N/A <	1.250e-01	0.125	ug/mL
4 SAMPLE	S98T002391	0 01C-01 NO3-02	LIQUID	N/A <	1.390e-01	0.139	ug/mL
4 SAMPLE	S98T002391	0 01C-01 P04-02	LIQUID	N/A <	1.200e-01	0.120	ug/mL
4 SAMPLE	S98T002391	0 01C-01 S04-02	LIQUID	N/A <	1.380e-01	0.138	ug/mL
4 SAMPLE	S98T002391	0 01C-01 OXALATE2	LIQUID	N/A <	1.050e-01	0.105	ug/mL
5 DUP	S98T002391	0 01C-01 F-02	LIQUID	<1.20e-2	<1.20e-2		RPD
5 DUP	S98T002391	0 01C-01 CL-02	LIQUID	3.30e-02	3.30e-02	0.000	RPD
5 DUP	S98T002391	0 01C-01 NO2-02	LIQUID	<1.08e-1	<1.08e-1		RPD
5 DUP	S98T002391	0 01C-01 BR-02	LIQUID	<1.25e-1	<1.25e-1		RPD
5 DUP	S98T002391	0 01C-01 NO3-02	LIQUID	<1.39e-1	<1.39e-1		RPD
5 DUP	S98T002391	0 01C-01 P04-02	LIQUID	<1.20e-1	<1.20e-1		RPD
5 DUP	S98T002391	0 01C-01 S04-02	LIQUID	<1.38e-1	<1.38e-1		RPD
5 DUP	S98T002391	0 01C-01 OXALATE2	LIQUID	<1.05e-1	<1.05e-1		RPD
6 SAMPLE	S98T002392	0 01C-01 F-02	LIQUID	N/A <	2.545e+01	25.450	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 25891

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002392	0	QIC-01	CL-02	LIQUID	N/A	< 3.606e+01	36.060 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	NO2-02	LIQUID	N/A	< 2.291e+02	229.100 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	BR-02	LIQUID	N/A	< 2.427e+04	265.100 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	NO3-02	LIQUID	N/A	< 2.948e+02	294.800 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	PO4-02	LIQUID	N/A	< 2.545e+02	254.500 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	SO4-02	LIQUID	N/A	< 2.927e+02	292.700 ug/mL
6 SAMPLE	S98T002392	0	QIC-01	OXALATE	LIQUID	N/A	< 2.227e+02	222.700 ug/mL
7 DUP	S98T002392	0	QIC-01	F-02	LIQUID	<2.54e1	<2.55e1	RPD
7 DUP	S98T002392	0	QIC-01	CL-02	LIQUID	<3.61e1	5.64e+01	RPD
7 DUP	S98T002392	0	QIC-01	NO2-02	LIQUID	<2.29e2	<2.29e2	RPD
7 DUP	S98T002392	0	QIC-01	BR-02	LIQUID	2.43e+04	2.40e+04	1.242 RPD
7 DUP	S98T002392	0	QIC-01	NO3-02	LIQUID	<2.95e2	<2.95e2	RPD
7 DUP	S98T002392	0	QIC-01	PO4-02	LIQUID	<2.54e2	<2.55e2	RPD
7 DUP	S98T002392	0	QIC-01	SO4-02	LIQUID	<2.93e2	<2.93e2	RPD
7 DUP	S98T002392	0	QIC-01	OXALATE	LIQUID	<2.23e2	<2.23e2	RPD

Final page for worklist# 25891

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 9/3/98

08/27/98 09:57
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 25891

Analyst: ADP Instrument: IC Book# 2/N21-B
 Method: LA-533-105 Rev/Mod F-0
 Worklist Comment: BY-105 POR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 ICV			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 SAMPLE	S98T002391 0		@IC-01	LIQUID	98000436	BY-105 (2)
	Analytes Requested:		BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02			
5 DUP	S98T002391 0		@IC-01	LIQUID		
6 SAMPLE	S98T002392 0		@IC-01	LIQUID	98000436	BY-105 (2)
	Analytes Requested:		BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02			
7 DUP	S98T002392 0		@IC-01	LIQUID		

Final page for worklist # 25891

Anthony P. Smith 9-3-98
 Analyst Signature Date

_____
 Analyst Signature Date

Validated 9/3/98 JML

Data Entry Comments:

up loaded 9-3-98
John W. Smith
25891/5EPCSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1098

Sample Name: 21N21-B Date: 09/03/1998 00:24:19
 Data File : C:\DX\DATA\98090201.D33
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: System: 1 Inject#: 33 Detector: CDM-1
 Analyst : *Anthony P. Smith* Column: AG4A/AS4A anion column

9-3-98

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	85	305	2	
2	0.99	fluoride	58.847	99.74	11757	2	-0.67
3	1.47	chloride	72.117	90.15	9600	1	-2.22
4	1.75	nitrite	512.270	95.75	48035	1	-3.50
5	2.57	bromide	538.389	91.88	33442	1	-1.53
6	2.91	nitrate	530.055	89.54	43122	1	-1.47
7	4.24	phosphate	476.332	87.08	18728	1	-0.62
8	5.44	sulfate	588.305	92.21	62069	1	-2.86
9	7.09	oxalate	532.364	95.58	43050	1	-2.21
Totals			3308.679	35651	270106		

File: 98090201.D33 Sample: 21N21-B

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1089 TO 1107.

SS-10-10

1099

Sample Name: 20N21-B Date: 09/03/1998 00:40:26
 Data File : C:\DX\DATA\98090201.D34
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 34 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00	30	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	93	345	2	
2	0.99	fluoride	60.657	2898	12126	2	-0.67
3	1.47	chloride	87.973	2790	11742	1	-2.22
4	1.74	nitrite	517.906	9668	48579	1	-3.87
5	2.57	bromide	602.576	6516	37585	1	-1.79
6	2.90	nitrate	681.620	8148	55926	1	-1.69
7	4.23	phosphate	578.774	1890	22899	1	-0.78
8	5.44	sulfate	667.640	5560	70448	1	-2.86
9	7.09	oxalate	518.503	2663	41913	1	-2.21
Totals			3715.651	40227	301563		

File: 98090201.D34 Sample: 20N21-B

35 106-10

1100

```

=====
Sample Name: BLANK                      Date: 09/03/1998 00:04:33
Data File  : C:\DX\DATA\98090201.D32
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1   System: 1   Inject#: 32   Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	14	48	1	
2	1.37		0.000	19	73	2	
3	1.47	chloride	0.014	52	203	2	-1.78
Totals			0.014	85	325		

File: 98090201.D32 Sample: BLANK

591

1101

=====

Sample Name: S98T002391 SAM	Date: 09/03/1998 02:27:14
Data File : C:\DX\DATA\98090201.D37	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 37	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.29		0.000	52	489	1	
2	1.49	chloride	0.033	105	451	1	-0.89
3	5.49	sulfate	-0.036	46	675	1	-1.90
Totals			-0.003	203	1615		

File: 98090201.D37 Sample: S98T002391 SAM

SS1

1102

Sample Name: S98T002391 DUP Date: 09/03/1998 02:45:24
 Data File : C:\DX\DATA\98090201.D38
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 38 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.30		0.000	49	446	1	
2	1.50	chloride	0.033	110	460	1	0.00
3	5.55	sulfate	-0.033	56	703	1	-0.95
Totals			0.000	215	1608		

SS1

1103


```

=====
Sample Name: S98T002392 SAM                      Date: 09/03/1998 03:23:14
Data File  : C:\DX\DATA\98090201.D41
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 41              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====
  
```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 2121 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	36	167	1	
2	1.48	chloride	31.417	22	212	1	-1.33
3	2.57	bromide	24269.657	12263	73980	1	-1.53
Totals			24301.073	12321	74359		

File: 98090201.D41 Sample: S98T002392 SAM

SS .100-10-.500-10

1104

```

=====
Sample Name: S98T002392 DUP                      Date: 09/03/1998 03:35:08
Data File  : C:\DX\DATA\98090201.D42
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 42              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          2121   3000 5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	48	264	1	
2	1.47	chloride	56.408	38	369	1	-1.78
3	2.57	bromide	23979.031	12281	73053	1	-1.53
Totals			24035.439	12367	73686		

File: 98090201.D42 Sample: S98T002392 DUP

SS . 100-10-500-10

LABCORE Completed Worklist Report for Worklist# 26756

Analyst: adp

Instrument: IC40S1

Book#: 28N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC-QC	F	QC	1	<1.20e-2		ug/mL
1 CCB	0	@IC-QC	CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	@IC-QC	NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	@IC-QC	BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	@IC-QC	NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	@IC-QC	P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	@IC-QC	S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	@IC-QC	OXALATEZ	QC	1	<1.05e-1		ug/mL
2 ICV	0	@IC-QC	F	QC	5.90e1	6.07e+01	102.881 % Recovery	
2 ICV	0	@IC-QC	CL	QC	8.00e1	8.00e+01	100.000 % Recovery	
2 ICV	0	@IC-QC	NO2	QC	5.70e2	5.71e+02	100.175 % Recovery	
2 ICV	0	@IC-QC	BR	QC	5.86e2	5.74e+02	97.952 % Recovery	
2 ICV	0	@IC-QC	NO3	QC	5.92e2	5.60e+02	94.595 % Recovery	
2 ICV	0	@IC-QC	P04	QC	5.47e2	5.36e+02	97.989 % Recovery	
2 ICV	0	@IC-QC	S04	QC	6.38e2	6.33e+02	99.216 % Recovery	
2 ICV	0	@IC-QC	OXALATEZ	QC	5.44e2	5.50e+02	101.103 % Recovery	
3 CCV	0	@IC-QC	F	QC	6.40e1	6.68e+01	104.375 % Recovery	
3 CCV	0	@IC-QC	CL	QC	9.00e1	9.18e+01	102.000 % Recovery	
3 CCV	0	@IC-QC	NO2	QC	5.53e2	5.67e+02	102.532 % Recovery	
3 CCV	0	@IC-QC	BR	QC	6.30e2	6.27e+02	99.524 % Recovery	
3 CCV	0	@IC-QC	NO3	QC	6.98e2	6.90e+02	98.854 % Recovery	
3 CCV	0	@IC-QC	P04	QC	6.32e2	6.29e+02	99.525 % Recovery	
3 CCV	0	@IC-QC	S04	QC	6.99e2	7.02e+02	100.429 % Recovery	
3 CCV	0	@IC-QC	OXALATEZ	QC	5.47e2	5.59e+02	102.194 % Recovery	
4 SAMPLE	S98T002814	0	@IC-01	F-02	LIQUID	N/A	5.183e+02	61.810 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	CL-02	LIQUID	N/A	3.991e+03	87.570 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	NO2-02	LIQUID	N/A	5.661e+04	556.300 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	BR-02	LIQUID	N/A	1.121e+04	643.900 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	NO3-02	LIQUID	N/A	2.554e+05	716.000 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	P04-02	LIQUID	N/A	1.048e+03	618.100 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	S04-02	LIQUID	N/A	2.141e+03	710.800 ug/mL
4 SAMPLE	S98T002814	0	@IC-01	OXALATEZ	LIQUID	N/A	5.409e+02	540.900 ug/mL
5 DUP	S98T002814	0	@IC-01	F-02	LIQUID	5.18e+02	4.97e+02	4.138 RPD
5 DUP	S98T002814	0	@IC-01	CL-02	LIQUID	3.99e+03	3.99e+03	0.000 RPD
5 DUP	S98T002814	0	@IC-01	NO2-02	LIQUID	5.66e+04	5.57e+04	1.603 RPD
5 DUP	S98T002814	0	@IC-01	BR-02	LIQUID	1.12e+03	1.15e+03	2.643 RPD
5 DUP	S98T002814	0	@IC-01	NO3-02	LIQUID	2.55e+05	2.51e+05	1.581 RPD
5 DUP	S98T002814	0	@IC-01	P04-02	LIQUID	1.05e+03	1.12e+03	6.452 RPD
5 DUP	S98T002814	0	@IC-01	S04-02	LIQUID	2.14e+03	2.17e+03	1.392 RPD
5 DUP	S98T002814	0	@IC-01	OXALATEZ	LIQUID	<5.41e2		RPD
6 SAMPLE	S98T002815	0	@IC-01	F-02	LIQUID	N/A	4.014e+02	61.810 ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26756

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002815	0	@IC-01 CL-02	LIQUID	N/A	5.451e+03	87.570	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 NO2-02	LIQUID	N/A	6.222e+04	556.300	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 BR-02	LIQUID	N/A	8.408e+02	643.900	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 NO3-02	LIQUID	N/A	2.232e+05	716.000	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 PO4-02	LIQUID	N/A	8.268e+02	618.100	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 SO4-02	LIQUID	N/A	2.184e+03	710.800	ug/mL
6 SAMPLE	S98T002815	0	@IC-01 OXALATE2	LIQUID	N/A	5.409e+02	540.900	ug/mL
7 DUP	S98T002815	0	@IC-01 F-02	LIQUID	4.01e+02	3.77e+02	6.170	RPD
7 DUP	S98T002815	0	@IC-01 CL-02	LIQUID	5.45e+03	5.92e+03	8.267	RPD
7 DUP	S98T002815	0	@IC-01 NO2-02	LIQUID	6.22e+04	6.42e+04	3.165	RPD
7 DUP	S98T002815	0	@IC-01 BR-02	LIQUID	8.41e+02	8.68e+02	3.160	RPD
7 DUP	S98T002815	0	@IC-01 NO3-02	LIQUID	2.23e+05	2.25e+05	0.893	RPD
7 DUP	S98T002815	0	@IC-01 PO4-02	LIQUID	8.27e+02	8.28e+02	0.121	RPD
7 DUP	S98T002815	0	@IC-01 SO4-02	LIQUID	2.18e+03	2.26e+03	3.604	RPD
7 DUP	S98T002815	0	@IC-01 OXALATE2	LIQUID	<5.41e2	<5.41e2		RPD
8 SAMPLE	S98T002816	0	@IC-01 F-02	LIQUID	N/A	2.940e+02	61.810	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 CL-02	LIQUID	N/A	4.689e+03	87.570	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 NO2-02	LIQUID	N/A	6.640e+04	556.300	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 BR-02	LIQUID	N/A	7.154e+02	643.900	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 NO3-02	LIQUID	N/A	2.089e+05	716.000	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 PO4-02	LIQUID	N/A	6.181e+02	618.100	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 SO4-02	LIQUID	N/A	2.115e+03	710.800	ug/mL
8 SAMPLE	S98T002816	0	@IC-01 OXALATE2	LIQUID	N/A	5.409e+02	540.900	ug/mL
9 DUP	S98T002816	0	@IC-01 F-02	LIQUID	2.94e+02	2.97e+02	1.015	RPD
9 DUP	S98T002816	0	@IC-01 CL-02	LIQUID	4.69e+03	4.69e+03	0.000	RPD
9 DUP	S98T002816	0	@IC-01 NO2-02	LIQUID	6.64e+04	6.60e+04	0.604	RPD
9 DUP	S98T002816	0	@IC-01 BR-02	LIQUID	7.15e+02	6.94e+02	2.981	RPD
9 DUP	S98T002816	0	@IC-01 NO3-02	LIQUID	2.09e+05	2.09e+05	0.000	RPD
9 DUP	S98T002816	0	@IC-01 PO4-02	LIQUID	<6.18e2	7.57e+02		RPD
9 DUP	S98T002816	0	@IC-01 SO4-02	LIQUID	2.12e+03	2.07e+03	2.387	RPD
9 DUP	S98T002816	0	@IC-01 OXALATE2	LIQUID	<5.41e2	<5.41e2		RPD

Final page for worklist# 26756

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John M. Luy
Reviewer Signature Date 10/26/98

10/23/98 09:41
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26756

Analyst: AD Instrument: IC Book# 29N21-AMethod: LA-533-105 Rev/Moi F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 ICV			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 SAMPLE	S98T002814 0		@IC-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
5 DUP	S98T002814 0		@IC-01	LIQUID		
6 SAMPLE	S98T002815 0		@IC-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
7 DUP	S98T002815 0		@IC-01	LIQUID		
8 SAMPLE	S98T002816 0		@IC-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
9 DUP	S98T002816 0		@IC-01	LIQUID		

Final page for worklist # 26756

10-25-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 10-26-98 validated 10/26/98 JMD
JH Warell
267560CT.CSV

S = Worklist Slot Number, R = Replica's Number, A = Aliquot Code.

1108

Sample Name: 29N21-A ICV Date: 10/25/1998 12:55:19
 Data File : C:\DX\DATA\98102501.D14
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 14 Detector: CDM-1
 Analyst : *Arthur Fanta* Column: AG4A/AS4A anion column

10-25-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	64	244	2	
2	1.00	fluoride	60.709	102.90	10443	2	1.01
3	1.47	chloride	79.954	79.94	9882	1	-1.57
4	1.74	nitrite	570.758	100.13	48048	1	-2.79
5	2.57	bromide	574.477	98.03	34800	1	1.05
6	2.90	nitrate	560.462	94.67	45521	1	-0.68
7	4.17	phosphate	535.986	97.99	19118	1	-0.63
8	5.44	sulfate	633.272	99.26	61716	1	-0.91
9	7.04	oxalate	550.227	101.15	40208	1	-1.54
Totals			3565.846	35080	269979		

File: 98102501.D14 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1109 TO 1117.

1109

Sample Name: 28N21-A CCV Date: 10/25/1998 13:06:39
 Data File : C:\DX\DATA\98102501.D15
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 15 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.21		0.000	20	190	1	
2	0.87		0.000	75	312	2	
3	1.01	fluoride	66.799	2201	11505	2	2.36
4	1.49	chloride	91.820	2527	11390	1	0.22
5	1.77	nitrite	567.207	9219	47743	1	-0.93
6	2.61	bromide	626.706	6395	38062	1	2.62
7	2.93	nitrate	690.166	8199	56410	1	0.46
8	4.22	phosphate	628.785	1852	22538	1	0.48
9	5.49	sulfate	701.524	5744	68548	1	0.06
10	7.09	oxalate	568.905	2282	40850	1	-0.79
Totals			3931.912	38513	297549		

File: 98102501.D15 Sample: 28N21-A CCV

1110

=====

Sample Name: BLANK	Date: 10/25/1998 12:41:09
Data File : C:\DX\DATA\98102501.D13	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 13	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.38		0.000	74	586	1	
Totals			0.000	74	586		

File: 98102501.D13 Sample: BLANK

1111

=====

Sample Name: S98T002814 SAM	Date: 10/25/1998 14:24:05
Data File : C:\DX\DATA\98102501.D16	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 16	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 5151 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	319	1545	2	
2	1.01	fluoride	518.261	369	1736	2	2.36
3	1.10		0.000	404	3372	2	
4	1.47	chloride	3991.198	1950	9667	1	-1.57
5	1.75	nitrite	56609.420	17440	94795	1	-2.05
6	2.60	bromide	1120.920	158	806	1	2.36
7	2.82	nitrate	255402.946	52812	469736	1	-3.42
8	4.21	phosphate	1047.853	45	521	1	0.16
9	5.49	sulfate	2141.141	273	3590	1	0.06
Totals			320831.738	73770	585768		

File: 98102501.D16 Sample: S98T002814 SAM

1112

Sample Name: S98T002814 DUP Date: 10/25/1998 14:38:11
 Data File : C:\DX\DATA\98102501.D17
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 17 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	322	1566	2	
2	1.01	fluoride	497.393	361	1667	2	1.68
3	1.09		0.000	385	2993	2	
4	1.46	chloride	3992.526	1965	9670	1	-2.01
5	1.74	nitrite	55688.632	17032	93217	1	-2.79
6	2.58	bromide	1154.521	166	846	1	1.57
7	2.81	nitrate	251089.154	52716	460391	1	-3.88
8	3.71		0.000	28	229	1	
9	4.19	phosphate	1120.374	47	571	1	-0.16
10	5.44	sulfate	2166.707	247	3638	1	-0.91
11	8.43		0.000	16	324	1	
Totals			315709.307	73286	575112		

File: 98102501.D17 Sample: S98T002814 DUP

1113

Sample Name: S98T002815 SAM Date: 10/25/1998 14:49:15
 Data File : C:\DX\DATA\98102501.D18
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 5151 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	369	2000	2	
2	1.02	fluoride	401.428	319	1346	2	3.03
3	1.09		0.000	418	3545	2	
4	1.46	chloride	5450.951	2656	13315	1	-2.01
5	1.75	nitrite	62217.448	19150	104430	1	-2.42
6	2.59	bromide	840.815	94	472	1	2.10
7	2.82	nitrate	223208.956	46885	401526	1	-3.42
8	3.71		0.000	30	258	1	
9	4.22	phosphate	826.793	35	367	1	0.48
10	5.44	sulfate	2184.154	263	3671	1	-0.91
11	9.17		0.000	19	575	1	
Totals			295130.546	70241	531507		

File: 98102501.D18 Sample: S98T002815 SAM

1114

```

=====
Sample Name: S98T002815  DUP                               Date: 10/25/1998 15:01:08
Data File   : C:\DX\DATA\98102501.D19
Method      : C:\DX\METHOD\KIT.MET
ACI Address : 1 System: 1 Inject#: 19                      Detector: CDM-1
Analyst     :                                               Column: AG4A/AS4A anion column
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          5151    3000  5Hz   0.00 10.00        30
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	369	2021	2	
2	1.02	fluoride	377.394	298	1266	2	3.03
3	1.09		0.000	423	3532	2	
4	1.46	chloride	5924.694	2834	14508	1	-2.01
5	1.75	nitrite	64158.629	19312	107774	1	-2.42
6	2.59	bromide	868.375	101	505	1	1.84
7	2.82	nitrate	225230.583	47054	405710	1	-3.42
8	3.69		0.000	63	723	1	
9	4.17	phosphate	827.945	38	368	1	-0.63
10	5.44	sulfate	2263.850	273	3823	1	-0.91
Totals			299651.471	70764	540230		

File: 98102501.D19 Sample: S98T002815 DUP

1115

Sample Name: S98T002816 SAM Date: 10/25/1998 15:12:54
 Data File : C:\DX\DATA\98102501.D20
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 20 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

PK. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	368	2035	2	
2	1.01	fluoride	293.973	245	988	2	2.36
3	1.09		0.000	423	3580	2	
4	1.45	chloride	4689.100	2277	11406	1	-2.46
5	1.74	nitrite	66399.001	20218	111637	1	-2.79
6	2.58	bromide	715.442	63	323	1	1.57
7	2.82	nitrate	208897.585	44379	372269	1	-3.42
8	3.71		0.000	44	417	1	
9	5.44	sulfate	2114.837	280	3540	1	-0.91
Totals			283109.937	68297	506195		

File: 98102501.D20 Sample: S98T002816 SAM

1116

Sample Name: S98T002816 DUP Date: 10/25/1998 15:23:43
 Data File : C:\DX\DATA\98102501.D21
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 21 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	373	1983	2	
2	1.02	fluoride	296.990	257	998	2	3.03
3	1.10		0.000	434	3712	2	
4	1.46	chloride	4694.742	2313	11420	1	-2.01
5	1.74	nitrite	65978.650	20120	110912	1	-2.79
6	2.59	bromide	693.509	61	297	1	1.84
7	2.83	nitrate	208805.836	44306	372083	1	-3.20
8	3.71		0.000	29	226	1	
9	4.19	phosphate	756.691	29	319	1	-0.16
10	4.77		0.000	60	1501	1	
11	5.44	sulfate	2074.246	265	3463	1	-0.91
Totals			283300.664	68247	506913		

File: 98102501.D21 Sample: S98T002816 DUP

1117

LABCORE Completed Worklist Report for Worklist# 26757

Analyst: adp

Instrument: IC40S1

Book#: 29N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	@IC-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 ICV	0	@IC-QC F	QC	5.90e1	6.31e+01	106.949 % Recovery	
2 ICV	0	@IC-QC CL	QC	8.00e1	8.18e+01	102.250 % Recovery	
2 ICV	0	@IC-QC NO2	QC	5.70e2	5.97e+02	104.737 % Recovery	
2 ICV	0	@IC-QC BR	QC	5.86e2	5.83e+02	99.488 % Recovery	
2 ICV	0	@IC-QC NO3	QC	5.92e2	5.84e+02	98.549 % Recovery	
2 ICV	0	@IC-QC P04	QC	5.47e2	5.50e+02	100.548 % Recovery	
2 ICV	0	@IC-QC S04	QC	6.38e2	6.52e+02	102.194 % Recovery	
2 ICV	0	@IC-QC OXALATE2	QC	5.44e2	5.64e+02	103.676 % Recovery	
3 CCV	0	@IC-QC F	QC	6.40e1	6.60e+01	103.125 % Recovery	
3 CCV	0	@IC-QC CL	QC	9.00e1	9.08e+01	100.889 % Recovery	
3 CCV	0	@IC-QC NO2	QC	5.43e2	5.55e+02	104.052 % Recovery	
3 CCV	0	@IC-QC BR	QC	6.30e2	6.11e+02	96.984 % Recovery	
3 CCV	0	@IC-QC NO3	QC	6.98e2	6.85e+02	98.138 % Recovery	
3 CCV	0	@IC-QC P04	QC	6.32e2	6.35e+02	100.475 % Recovery	
3 CCV	0	@IC-QC S04	QC	6.99e2	7.09e+02	101.431 % Recovery	
3 CCV	0	@IC-QC OXALATE2	QC	5.28e2	5.63e+02	106.629 % Recovery	
4 SAMPLE	S98T002817 0	@IC-01 F-02	LIQUID	N/A	4.181e+02		61.810 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 CL-02	LIQUID	N/A	4.556e+03		87.570 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 NO2-02	LIQUID	N/A	6.476e+04		556.300 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 BR-02	LIQUID	N/A	6.472e+02		643.900 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 NO3-02	LIQUID	N/A	2.142e+03		716.000 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 P04-02	LIQUID	N/A	7.573e+02		618.100 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 S04-02	LIQUID	N/A	1.972e+03		710.800 ug/mL
4 SAMPLE	S98T002817 0	@IC-01 OXALATE2	LIQUID	N/A	< 5.409e+02		540.900 ug/mL
5 DUP	S98T002817 0	@IC-01 F-02	LIQUID	4.18e+02	3.95e+02		5.658 RPD
5 DUP	S98T002817 0	@IC-01 CL-02	LIQUID	4.56e+03	4.44e+03		2.667 RPD
5 DUP	S98T002817 0	@IC-01 NO2-02	LIQUID	6.48e+04	6.36e+04		1.869 RPD
5 DUP	S98T002817 0	@IC-01 BR-02	LIQUID	6.47e+02	6.54e+02		1.076 RPD
5 DUP	S98T002817 0	@IC-01 NO3-02	LIQUID	2.14e+05	2.19e+05		2.309 RPD
5 DUP	S98T002817 0	@IC-01 P04-02	LIQUID	7.57e+02	2.77e+03		114.148 RPD
5 DUP	S98T002817 0	@IC-01 S04-02	LIQUID	1.97e+03	1.98e+03		0.506 RPD
5 DUP	S98T002817 0	@IC-01 OXALATE2	LIQUID	<5.41e2	<5.41e2		RPD
6 SAMPLE	S98T002818 0	@IC-01 F-02	LIQUID	N/A	6.181e+01		61.810 ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26757

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002818	0	①C-01	CL-02	LIQUID	N/A	5.037e+03	87.570 ug/mL
6 SAMPLE	S98T002818	0	①C-01	NO2-02	LIQUID	N/A	7.088e+04	556.300 ug/mL
6 SAMPLE	S98T002818	0	①C-01	BR-02	LIQUID	N/A	5.439e+02	643.900 ug/mL
6 SAMPLE	S98T002818	0	①C-01	NO3-02	LIQUID	N/A	1.881e+05	716.000 ug/mL
6 SAMPLE	S98T002818	0	①C-01	PO4-02	LIQUID	N/A	1.528e+03	618.100 ug/mL
6 SAMPLE	S98T002818	0	①C-01	SO4-02	LIQUID	N/A	2.320e+03	710.800 ug/mL
6 SAMPLE	S98T002818	0	①C-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
7 DUP	S98T002818	0	①C-01	F-02	LIQUID	<6.18e1	<6.18e1	RPD
7 DUP	S98T002818	0	①C-01	CL-02	LIQUID	5.04e+03	5.23e+03	3.700 RPD
7 DUP	S98T002818	0	①C-01	NO2-02	LIQUID	7.09e+04	7.34e+04	3.465 RPD
7 DUP	S98T002818	0	①C-01	BR-02	LIQUID	<6.44e2	<6.44e2	RPD
7 DUP	S98T002818	0	①C-01	NO3-02	LIQUID	1.88e+05	1.90e+05	1.058 RPD
7 DUP	S98T002818	0	①C-01	PO4-02	LIQUID	1.53e+03	<6.18e2	RPD
7 DUP	S98T002818	0	①C-01	SO4-02	LIQUID	2.32e+03	2.24e+03	3.509 RPD
7 DUP	S98T002818	0	①C-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD
8 SAMPLE	S98T002819	0	①C-01	F-02	LIQUID	N/A	6.181e+01	61.810 ug/mL
8 SAMPLE	S98T002819	0	①C-01	CL-02	LIQUID	N/A	5.142e+03	87.570 ug/mL
8 SAMPLE	S98T002819	0	①C-01	NO2-02	LIQUID	N/A	7.449e+04	556.300 ug/mL
8 SAMPLE	S98T002819	0	①C-01	BR-02	LIQUID	N/A	5.439e+02	643.900 ug/mL
8 SAMPLE	S98T002819	0	①C-01	NO3-02	LIQUID	N/A	1.865e+05	716.000 ug/mL
8 SAMPLE	S98T002819	0	①C-01	PO4-02	LIQUID	N/A	8.697e+02	618.100 ug/mL
8 SAMPLE	S98T002819	0	①C-01	SO4-02	LIQUID	N/A	2.128e+03	710.800 ug/mL
8 SAMPLE	S98T002819	0	①C-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
9 DUP	S98T002819	0	①C-01	F-02	LIQUID	<6.18e1	<6.18e1	RPD
9 DUP	S98T002819	0	①C-01	CL-02	LIQUID	5.14e+03	5.19e+03	0.966 RPD
9 DUP	S98T002819	0	①C-01	NO2-02	LIQUID	7.45e+04	7.45e+04	0.000 RPD
9 DUP	S98T002819	0	①C-01	BR-02	LIQUID	<6.44e2	<6.44e2	RPD
9 DUP	S98T002819	0	①C-01	NO3-02	LIQUID	1.86e+05	1.86e+05	0.000 RPD
9 DUP	S98T002819	0	①C-01	PO4-02	LIQUID	8.70e+02	1.07e+03	20.619 RPD
9 DUP	S98T002819	0	①C-01	SO4-02	LIQUID	2.13e+03	2.07e+03	2.857 RPD
9 DUP	S98T002819	0	①C-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD
10 SPK	S98T002819	0	①C-01	F-02	LIQUID	5.90e1	6.96e+01	117.966 % Recovery
10 SPK	S98T002819	0	①C-01	CL-02	LIQUID	8.00e+01	7.97e+01	99.625 % Recovery
10 SPK	S98T002819	0	①C-01	NO2-02	LIQUID	5.70e2	5.86e+02	102.807 % Recovery
10 SPK	S98T002819	0	①C-01	BR-02	LIQUID	5.86e2	5.39e+02	91.980 % Recovery
10 SPK	S98T002819	0	①C-01	NO3-02	LIQUID	5.92e2	5.41e+02	91.385 % Recovery
10 SPK	S98T002819	0	①C-01	PO4-02	LIQUID	5.47e2	5.06e+02	92.505 % Recovery
10 SPK	S98T002819	0	①C-01	SO4-02	LIQUID	6.38e2	6.33e+02	99.216 % Recovery
10 SPK	S98T002819	0	①C-01	OXALATE2	LIQUID	5.44e2	5.43e+02	99.816 % Recovery

Final page for worklist# 26757

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John M. Lutz
Reviewer Signature _____ Date 10/27/98

Units shown for QC (BLK/BKG) may not reflect the actual units.

10/23/98 09:41
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26757

Analyst: ASD Instrument: ICBook# 29N21-AMethod: LA-533-105 Rev/Mod F-0

Worklist: Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 ICV			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 SAMPLE	S98T002817 0	@IC-01	LIQUID		98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
5 DUP	S98T002817 0	@IC-01	LIQUID			
6 SAMPLE	S98T002818 0	@IC-01	LIQUID		98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
7 DUP	S98T002818 0	@IC-01	LIQUID			
8 SAMPLE	S98T002819 0	@IC-01	LIQUID		98000437	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
9 DUP	S98T002819 0	@IC-01	LIQUID			
10 SPK	S98T002819 0	@IC-01	LIQUID			

Final page for worklist # 26757

10-24-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 10-27-98
John Worsell

Validated 10/27/98 mby

26757OCT.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-A ICV Date: 10/24/1998 14:42:36
 Data File : C:\DX\DATA\98102411.D15
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 15 Detector: CDM-1
 Analyst : *John P. Smith* Column: AG4A/AS4A anion column

10-24-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	78	333	2	
2	0.99	fluoride	63.081/06.92	2217	10856	2	0.34
3	1.47	chloride	81.818/102.27	2284	10118	1	-1.57
4	1.75	nitrite	596.586/104.66	9671	50269	1	-2.42
5	2.57	bromide	582.724/99.44	6041	35315	1	1.31
6	2.90	nitrate	584.154/98.67	6832	47505	1	-0.68
7	4.18	phosphate	550.214/100.59	1605	19641	1	-0.48
8	5.44	sulfate	651.824/102.17	5259	63571	1	-0.91
9	7.09	oxalate	563.943/103.67	2553	41223	1	-0.79
Totals			3674.345	36541	278829		

File: 98102411.D15 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1121 TO 1130.

SS .100-10

1121

Sample Name: 28N21-A CCV Date: 10/24/1998 15:21:21
 Data File : C:\DX\DATA\98102411.D18
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	70	251	2	
2	0.99	fluoride	66.038	2278	11372	2	0.34
3	1.47	chloride	90.781	2562	11258	1	-1.57
4	1.73	nitrite	565.055	9285	47558	1	-3.17
5	2.56	bromide	610.922	6528	37075	1	0.79
6	2.89	nitrate	685.260	8291	55997	1	-1.14
7	4.17	phosphate	634.829	1884	22762	1	-0.63
8	5.44	sulfate	708.816	5785	69280	1	-0.91
9	7.04	oxalate	562.685	2312	41130	1	-1.54
Totals			3924.386	38995	296682		

File: 98102411.D18 Sample: 28N21-A CCV

SS .100-10

1122

Sample Name: BLANK

Date: 10/24/1998 14:31:36

Data File : C:\DX\DATA\98102411.D14

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 14

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.05	fluoride	0.002	12	39	1	5.72
2	1.36		0.000	30	192	1	
Totals			0.002	42	232		

File: 98102411.D14 Sample: BLANK

SSI

Sample Name: S98T002817 SAM Date: 10/24/1998 16:00:58
 Data File : C:\DX\DATA\98102411.D21
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 21 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	32	300	1	
2	0.89		0.000	330	1745	2	
3	1.05	fluoride	418.133	309	1402	2	5.72
4	1.13		0.000	420	3390	2	
5	1.51	chloride	4556.325	2233	11075	1	1.12
6	1.80	nitrite	64756.516	20030	108804	1	0.56
7	2.63	bromide	647.203	51	242	1	3.67
8	2.85	nitrate	214198.544	46815	383034	1	-2.28
9	3.77		0.000	29	231	1	
10	4.23	phosphate	757.267	33	319	1	0.79
11	5.49	sulfate	1972.286	265	3269	1	0.06
12	7.15	oxalate	371.336	22	439	1	-0.05
Totals			287677.610	70569	514249		

File: 98102411.D21 Sample: S98T002817 SAM

SS 100-10-.200-10

1124

```

=====
Sample Name: S98T002817 DUP                      Date: 10/24/1998 16:12:15
Data File  : F:\DATA\98102411.D22
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 22              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	349	1786	2	
2	1.01	fluoride	395.468	314	1326	2	1.68
3	1.09		0.000	434	3412	2	
4	1.45	chloride	4443.022	2241	10792	1	-2.46
5	1.73	nitrite	63642.933	19959	106885	1	-3.17
6	2.55	bromide	653.926	51	250	1	0.52
7	2.79	nitrate	218692.346	47313	392226	1	0.12
8	3.73		0.000	25	184	1	
9	4.20	phosphate	2772.664	85	1721	1	0.00
10	5.44	sulfate	1979.193	237	3282	1	-0.91
Totals			292579.552	71009	521864		

File: 98102411.D22 Sample: S98T002817 DUP

Data Reprocessed On 10/26/1998 15:47:01

Sample Name: S98T002818 SAM	Date: 10/24/1998 16:23:10
Data File : F:\DATA\98102411.D23	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 23	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	413	2292	2	
2	1.09		0.000	453	4587	2	
3	1.44	chloride	5036.687	2519	12276	1	-3.36
4	1.73	nitrite	70884.857	22415	119391	1	-3.54
5	2.55	bromide	502.307	16	69	1	0.26
6	2.79	nitrate	188111.865	40895	330830	1	-0.12
7	3.71		0.000	35	283	1	
8	4.56	phosphate	1527.745	35	855	1	8.57
9	5.44	sulfate	2320.033	305	3929	1	-0.91
10	7.15	oxalate	259.368	16	280	1	-0.05
Totals			268642.864	67103	474792		

File: 98102411.D23 Sample: S98T002818 SAM

Data Reprocessed On 10/26/1998 15:47:44

```

=====
Sample Name: S98T002818 DUP                      Date: 10/24/1998 16:34:45
Data File  : F:\DATA\98102411.D24
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 24              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      5151    3000    5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	415	2356	2	
2	1.09		0.000	447	4481	2	
3	1.45	chloride	5234.781	2544	12772	1	-2.46
4	1.73	nitrite	73441.516	22373	123821	1	-3.17
5	2.55	bromide	510.040	17	79	1	0.52
6	2.80	nitrate	189719.138	41092	333992	1	0.00
7	3.75		0.000	44	432	1	
8	5.44	sulfate	2238.287	288	3774	1	-0.91
Totals			271143.762	67220	481706		

File: 98102411.D24 Sample: S98T002818 DUP

Data Reprocessed On 10/26/1998 15:48:29

```

=====
Sample Name: S98T002819 SAM                      Date: 10/24/1998 16:45:42
Data File  : F:\DATA\98102411.D25
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 25              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	428	2363	2	
2	1.09		0.000	460	4695	2	
3	1.45	chloride	5141.974	2535	12540	1	-2.91
4	1.73	nitrite	74487.919	23201	125636	1	-3.17
5	2.80	nitrate	186489.238	40640	327645	1	0.00
6	3.73		0.000	32	235	1	
7	4.20	phosphate	869.686	43	397	1	0.00
8	5.44	sulfate	2128.437	271	3566	1	-0.91
Totals			269117.254	67609	477076		

File: 98102411.D25 Sample: S98T002819 SAM

```

=====
Sample Name: S98T002819 DUP                      Date: 10/24/1998 16:56:23
Data File  : F:\DATA\98102411.D26
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 26              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000    5Hz    0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	435	2487	2	
2	1.09		0.000	459	4563	2	
3	1.45	chloride	5186.034	2573	12650	1	-2.91
4	1.73	nitrite	74527.279	23378	125704	1	-3.17
5	2.80	nitrate	185791.814	40835	326278	1	0.00
6	3.74		0.000	32	231	1	
7	4.19	phosphate	1065.121	52	533	1	-0.32
8	4.75		0.000	53	1306	1	
9	5.44	sulfate	2069.607	263	3454	1	-0.91
Totals			268639.855	68080	477206		

File: 98102411.D26 Sample: S98T002819 DUP

```

=====
Sample Name: S98T002819 SPK                      Date: 10/24/1998 17:07:08
Data File  : F:\DATA\98102411.D27
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 27              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	461	2263	2	
2	1.01	fluoride	3517.189	1568	11884	2	1.68
3	1.45	chloride	9167.094	4621	22796	1	-2.91
4	1.73	nitrite	104097.629	32395	177532	1	-3.17
5	2.54	bromide	27231.539	6051	32275	1	0.00
6	2.79	nitrate	213824.378	46586	382271	1	-0.12
7	3.73		0.000	38	326	1	
8	4.17	phosphate	26446.598	1589	18478	1	-0.63
9	5.44	sulfate	34106.911	5440	65266	1	-0.91
10	7.04	oxalate	27439.833	2336	39306	1	-1.54
Totals			445831.171	101085	752396		

File: 98102411.D27 Sample: S98T002819 SPK

LABCORE Completed Worklist Report for Worklist# 26758

Analyst: adp

Instrument: IC40S1

Book#: 29N21A

Method: LA-533-105 Rev/Mod F 0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC PO4	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC SO4	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	ICV	0	@IC-QC F	QC	5.90e1	5.30e+01	105.780 % Recovery	
2	ICV	0	@IC-QC CL	QC	8.00e1	8.45e+01	105.625 % Recovery	
2	ICV	0	@IC-QC NO2	QC	5.70e2	5.78e+02	101.404 % Recovery	
2	ICV	0	@IC-QC BR	QC	5.86e2	5.70e+02	97.270 % Recovery	
2	ICV	0	@IC-QC NO3	QC	5.92e2	5.74e+02	96.959 % Recovery	
2	ICV	0	@IC-QC PO4	QC	5.47e2	5.71e+02	104.388 % Recovery	
2	ICV	0	@IC-QC SO4	QC	6.38e2	6.54e+02	102.508 % Recovery	
2	ICV	0	@IC-QC OXALATE2	QC	5.44e2	5.65e+02	103.860 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.47e+01	101.094 % Recovery	
3	CCV	0	@IC-QC CL	QC	9.00e1	9.47e+01	105.222 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.53e2	5.67e+02	102.532 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.27e+02	99.524 % Recovery	
3	CCV	0	@IC-QC NO3	QC	6.98e2	6.95e+02	99.570 % Recovery	
3	CCV	0	@IC-QC PO4	QC	6.32e2	6.57e+02	103.956 % Recovery	
3	CCV	0	@IC-QC SO4	QC	6.99e2	7.09e+02	101.431 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.47e2	5.84e+02	106.764 % Recovery	
4	SAMPLE	S98T002820	0	@IC-01 F-02	LIQUID	N/A	< 6.181e+01	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 CL-02	LIQUID	N/A	5.311e+03	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 NO2-02	LIQUID	N/A	7.512e+04	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 BR-02	LIQUID	N/A	< 6.439e+02	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 NO3-02	LIQUID	N/A	1.794e+05	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 PO4-02	LIQUID	N/A	1.423e+03	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 SO4-02	LIQUID	N/A	2.601e+03	ug/mL
4	SAMPLE	S98T002820	0	@IC-01 OXALATE2	LIQUID	N/A	9.877e+02	ug/mL
5	DUP	S98T002820	0	@IC-01 F-02	LIQUID	<6.18e1	<6.18e1	RPD
5	DUP	S98T002820	0	@IC-01 CL-02	LIQUID	5.31e+03	5.36e+03	0.937 RPD
5	DUP	S98T002820	0	@IC-01 NO2-02	LIQUID	7.51e+04	7.46e+04	0.668 RPD
5	DUP	S98T002820	0	@IC-01 BR-02	LIQUID	<6.44e2	<6.44e2	RPD
5	DUP	S98T002820	0	@IC-01 NO3-02	LIQUID	1.79e+05	1.79e+05	0.000 RPD
5	DUP	S98T002820	0	@IC-01 PO4-02	LIQUID	1.42e+03	1.39e+03	2.135 RPD
5	DUP	S98T002820	0	@IC-01 SO4-02	LIQUID	2.60e+03	2.54e+03	2.335 RPD
5	DUP	S98T002820	0	@IC-01 OXALATE2	LIQUID	9.88e+02	9.87e+02	50.921 RPD
6	SAMPLE	S98T002821	0	@IC-01 F-02	LIQUID	N/A	< 6.181e+01	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26758

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002821	0	①IC-01	CL-02	LIQUID	N/A	5.156e+03	87.570 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	NO2-02	LIQUID	N/A	7.948e+04	556.300 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	BR-02	LIQUID	N/A	6.439e+02	643.900 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	NO3-02	LIQUID	N/A	1.756e+05	716.000 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	PO4-02	LIQUID	N/A	1.453e+03	618.100 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	SO4-02	LIQUID	N/A	4.191e+03	710.800 ug/mL
6 SAMPLE	S98T002821	0	①IC-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
7 DUP	S98T002821	0	①IC-01	F-02	LIQUID	<6.18e1	<6.18e1	RPD
7 DUP	S98T002821	0	①IC-01	CL-02	LIQUID	5.16e+03	5.31e+03	2.865 RPD
7 DUP	S98T002821	0	①IC-01	NO2-02	LIQUID	7.95e+04	8.17e+04	2.730 RPD
7 DUP	S98T002821	0	①IC-01	BR-02	LIQUID	<6.44e2	<6.44e2	RPD
7 DUP	S98T002821	0	①IC-01	NO3-02	LIQUID	1.76e+05	1.74e+05	1.143 RPD
7 DUP	S98T002821	0	①IC-01	PO4-02	LIQUID	1.45e+03	3.67e+03	86.719 RPD
7 DUP	S98T002821	0	①IC-01	SO4-02	LIQUID	4.19e+03	4.02e+03	4.141 RPD
7 DUP	S98T002821	0	①IC-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD
8 SAMPLE	S98T002960	0	①IC-01	F-02	LIQUID	N/A	4.313e+02	61.810 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	CL-02	LIQUID	N/A	4.378e+03	87.570 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	NO2-02	LIQUID	N/A	6.206e+04	556.300 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	BR-02	LIQUID	N/A	6.439e+02	643.900 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	NO3-02	LIQUID	N/A	2.323e+05	716.000 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	PO4-02	LIQUID	N/A	8.426e+02	618.100 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	SO4-02	LIQUID	N/A	1.838e+03	710.800 ug/mL
8 SAMPLE	S98T002960	0	①IC-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
9 DUP	S98T002960	0	①IC-01	F-02	LIQUID	4.31e+02	4.36e+02	1.153 RPD
9 DUP	S98T002960	0	①IC-01	CL-02	LIQUID	4.38e+03	4.33e+03	1.148 RPD
9 DUP	S98T002960	0	①IC-01	NO2-02	LIQUID	6.21e+04	6.20e+04	0.161 RPD
9 DUP	S98T002960	0	①IC-01	BR-02	LIQUID	<6.44e2	<6.44e2	RPD
9 DUP	S98T002960	0	①IC-01	NO3-02	LIQUID	2.32e+05	2.32e+05	0.000 RPD
9 DUP	S98T002960	0	①IC-01	PO4-02	LIQUID	8.43e+02	9.12e+02	7.863 RPD
9 DUP	S98T002960	0	①IC-01	SO4-02	LIQUID	1.84e+03	1.84e+03	0.000 RPD
9 DUP	S98T002960	0	①IC-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD
10 SPK	S98T002960	0	①IC-01	F-02	LIQUID	5.90e1	7.49e+01	126.949 % Recovery
10 SPK	S98T002960	0	①IC-01	CL-02	LIQUID	8.00e1	7.74e+01	96.750 % Recovery
10 SPK	S98T002960	0	①IC-01	NO2-02	LIQUID	5.70e2	5.94e+02	104.211 % Recovery
10 SPK	S98T002960	0	①IC-01	BR-02	LIQUID	5.86e2	5.35e+02	91.297 % Recovery
10 SPK	S98T002960	0	①IC-01	NO3-02	LIQUID	5.32e2	4.81e+02	81.250 % Recovery
10 SPK	S98T002960	0	①IC-01	PO4-02	LIQUID	5.47e2	5.08e+02	92.870 % Recovery
10 SPK	S98T002960	0	①IC-01	SO4-02	LIQUID	6.38e2	6.29e+02	98.589 % Recovery
10 SPK	S98T002960	0	①IC-01	OXALATE2	LIQUID	5.44e2	5.58e+02	102.574 % Recovery
11 SAMPLE	S98T002961	0	①IC-01	F-02	LIQUID	N/A	4.094e+02	61.810 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	CL-02	LIQUID	N/A	4.245e+03	87.570 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	NO2-02	LIQUID	N/A	6.070e+04	556.300 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	BR-02	LIQUID	N/A	6.439e+02	643.900 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	NO3-02	LIQUID	N/A	2.080e+05	716.000 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	PO4-02	LIQUID	N/A	7.384e+02	618.100 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	SO4-02	LIQUID	N/A	1.685e+03	710.800 ug/mL
11 SAMPLE	S98T002961	0	①IC-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
12 DUP	S98T002961	0	①IC-01	F-02	LIQUID	4.09e+02	4.10e+02	0.244 RPD
12 DUP	S98T002961	0	①IC-01	CL-02	LIQUID	4.24e+03	4.29e+03	1.172 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 26758

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
12	DUP	S98T002961	0	@IC-01	NO2-02	LIQUID	6.07e+04	6.24e+04	2.762	RPD
12	DUP	S98T002961	0	@IC-01	BR-02	LIQUID	<6.44e2	<6.44e2		RPD
12	DUP	S98T002961	0	@IC-01	NO3-02	LIQUID	2.08e+05	2.09e+05	0.480	RPD
12	DUP	S98T002961	0	@IC-01	PO4-02	LIQUID	7.38e+02	6.90e+02	6.723	RPD
12	DUP	S98T002961	0	@IC-01	SO4-02	LIQUID	1.68e+03	1.80e+03	6.897	RPD
12	DUP	S98T002961	0	@IC-01	OXALATE2	LIQUID	<5.41e2	<5.41e2		RPD

Final page for worklist# 26758

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 10/27/98

LABCORE Data Entry Template for Worklist# 26758

Analyst: ADD Instrument: IC Book# 29N21-A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	ICV				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	SAMPLE	S98T002820 0			@IC-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
5	DUP	S98T002820 0			@IC-01	LIQUID		
6	SAMPLE	S98T002821 0			@IC-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
7	DUP	S98T002821 0			@IC-01	LIQUID		
8	SAMPLE	S98T002960 0			@IC-01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
9	DUP	S98T002960 0			@IC-01	LIQUID		
10	SPK	S98T002960 0			@IC-01	LIQUID		
11	SAMPLE	S98T002961 0			@IC-01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
12	DUP	S98T002961 0			@IC-01	LIQUID		

Final page for worklist # 26758

Anthony Pimentra 10-25-98

Data Entry Comments:

uploaded 10-27-98

validated 10/27/98 mgf

J. H. well
267580.CCBV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-A ICV Date: 10/25/1998 08:17:19
 Data File : C:\DX\DATA\98102501.D02
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 2 Detector: CDM-1
 Analyst: *Ally Kuruk* Column: AG4A/AS4A anion column

10-25-98
 Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	67	268	2	
2	1.00	fluoride	62.964	106.72	2159	2	1.01
3	1.49	chloride	84.489	105.61	2294	1	-0.22
4	1.77	nitrite	577.522	101.32	9386	1	-1.30
5	2.64	bromide	569.589	97.12	5822	1	3.94
6	2.98	nitrate	573.875	96.94	6558	1	2.05
7	4.15	phosphate	570.508	104.297	1587	1	-1.11
8	5.44	sulfate	654.228	102.54	5291	1	-0.91
9	7.09	oxalate	565.114	103.83	2543	1	-0.79

Totals 3658.289 35707 276838

File: 98102501.D02 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1135 TO 1146.

Sample Name: 28N21-A CCV Date: 10/25/1998 08:33:21
 Data File : C:\DX\DATA\98102501.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	92	394	2	
2	1.00	fluoride	64.698	2275	11138	2	1.01
3	1.49	chloride	94.730	2669	11761	1	-0.22
4	1.77	nitrite	566.738	9330	47703	1	-1.30
5	2.63	bromide	626.610	6401	38056	1	3.67
6	2.97	nitrate	695.486	8086	56858	1	1.60
7	4.15	phosphate	657.262	1877	23593	1	-1.11
8	5.39	sulfate	709.099	4539	69308	1	-1.88
9	7.04	oxalate	584.500	2380	42745	1	-1.54
Totals			3999.122	37651	301557		

File: 98102501.D03 Sample: 28N21-A CCV

Sample Name: BLANK Date: 10/25/1998 08:02:12
Data File : C:\DX\DATA\98102501.D01
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 1 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.05		0.000	158	320	1	
2	0.17		0.000	37	114	1	
3	0.64		0.000	34	89	1	
4	1.99		0.000	19	117	1	
Totals			0.000	248	640		

File: 98102501.D01 Sample: BLANK

```

=====
Sample Name: S98T002820 SAM                      Date: 10/25/1998 09:01:10
Data File  : C:\DX\DATA\98102501.D04
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 4                Detector: CDM-1
Analyst    :                                     Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151  3000  5Hz  0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	448	2549	2	
2	1.10		0.000	464	4819	2	
3	1.47	chloride	5310.834	2613	12963	1	-1.57
4	1.76	nitrite	75124.017	22962	126740	1	-1.68
5	2.88	nitrate	179364.703	37493	313742	1	-1.37
6	3.67		0.000	28	209	1	
7	4.17	phosphate	1422.903	65	782	1	-0.63
8	5.44	sulfate	2600.638	322	4462	1	-0.91
9	7.36	oxalate	987.700	37	1316	1	2.94
Totals			264810.796	64431	467581		

File: 98102501.D04 Sample: S98T002820 SAM

Sample Name: S98T002820 DUP Date: 10/25/1998 09:14:25
 Data File : C:\DX\DATA\98102501.D05
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 5 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	458	2661	2	
2	1.09		0.000	467	4852	2	
3	1.47	chloride	5355.593	2599	13076	1	-1.57
4	1.75	nitrite	74597.469	22855	125826	1	-2.05
5	2.87	nitrate	178935.346	37532	312909	1	-1.60
6	3.64		0.000	30	210	1	
7	4.17	phosphate	1389.103	65	758	1	-0.79
8	5.44	sulfate	2544.735	331	4356	1	-0.91
9	7.15	oxalate	587.087	23	746	1	-0.05

Totals 263409.333 64359 465393

File: 98102501.D05 Sample: S98T002820 DUP

Sample Name: S98T002821 SAM Date: 10/25/1998 10:53:53
 Data File : C:\DX\DATA\98102501.D06
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 6 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	475	2658	2	
2	1.09		0.000	421	4853	2	
3	1.45	chloride	5155.891	2548	12575	1	-2.91
4	1.73	nitrite	79475.648	24345	134305	1	-3.17
5	2.84	nitrate	175628.513	37451	306504	1	-2.74
6	3.65		0.000	19	122	1	
7	4.15	phosphate	1452.962	72	803	1	-1.11
8	5.39	sulfate	4191.039	490	7483	1	-1.88
Totals			265904.054	65822	469303		

File: 98102501.D06 Sample: S98T002821 SAM

```

=====
Sample Name: S98T002821 DUP                      Date: 10/25/1998 11:05:05
Data File  : C:\DX\DATA\98102501.D07
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 7                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000 5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	442	2594	2	
2	1.10		0.000	402	4517	2	
3	1.47	chloride	5305.019	2493	12949	1	-1.57
4	1.75	nitrite	81702.006	24282	138184	1	-2.05
5	2.86	nitrate	174407.869	37092	304147	1	-2.05
6	3.68		0.000	20	165	1	
7	4.20	phosphate	3668.583	105	2345	1	0.00
8	5.44	sulfate	4017.228	527	7152	1	-0.91
9	7.15	oxalate	263.237	18	285	1	-0.05
Totals			269363.942	65382	472339		

File: 98102501.D07 Sample: S98T002821 DUP

Sample Name: S98T002960 SAM Date: 10/25/1998 11:19:08
 Data File : C:\DX\DATA\98102501.D08
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 8 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	357	1849	2	
2	1.01	fluoride	431.263	327	1446	2	2.36
3	1.10		0.000	422	3422	2	
4	1.46	chloride	4378.371	2122	10631	1	-3.01
5	1.75	nitrite	62061.735	18964	104162	1	-2.42
6	2.60	bromide	544.080	25	119	1	2.36
7	2.83	nitrate	232253.421	48470	420342	1	-3.20
8	3.70		0.000	24	158	1	
9	4.18	phosphate	842.627	40	378	1	-0.48
10	5.44	sulfate	1838.151	222	3015	1	-0.91
11	7.15	oxalate	378.650	23	449	1	-0.05
Totals			302728.299	70996	545972		

File: 98102501.D08 Sample: S98T002960 SAM

```

=====
Sample Name: S98T002960 DUP                      Date: 10/25/1998 11:33:24
Data File  : C:\DX\DATA\98102501.D09
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 9                Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          5151  3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	349	1769	2	
2	1.01	fluoride	436.229	331	1462	2	2.36
3	1.10		0.000	435	3574	2	
4	1.46	chloride	4330.343	2125	10511	1	-2.01
5	1.75	nitrite	62014.495	19121	104081	1	-2.42
6	2.59	bromide	531.389	23	104	1	2.10
7	2.82	nitrate	232010.203	48577	419833	1	-3.42
8	3.71		0.000	24	168	1	
9	4.17	phosphate	911.569	40	426	1	-0.79
10	5.44	sulfate	1836.569	217	3012	1	-0.91
11	7.15	oxalate	330.826	20	381	1	-0.05
Totals			302401.625	71263	545321		

File: 98102501.D09 Sample: S98T002960 DUP

```

=====
Sample Name: S98T002960 SPK                      Date: 10/25/1998 11:44:08
Data File  : C:\DX\DATA\98102501.D10
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 10              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151  3000  5Hz  0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	404	1875	2	
2	1.01	fluoride	4212.955	2102	14277	2	1.68
3	1.45	chloride	8286.144	4080	20523	1	-2.46
4	1.75	nitrite	92044.879	28293	156281	1	-2.42
5	2.57	bromide	27018.067	6014	32015	1	1.31
6	2.81	nitrate	256569.914	53908	472275	1	-3.88
7	3.71		0.000	29	224	1	
8	4.17	phosphate	26489.823	1560	18509	1	-0.63
9	5.44	sulfate	33616.390	5333	64303	1	-0.91
10	7.04	oxalate	28162.341	2261	40354	1	-1.54
Totals			476400.513	103984	820636		

File: 98102501.D10 Sample: S98T002960 SPK

```

=====
Sample Name: S98T002961 SAM                      Date: 10/25/1998 11:55:03
Data File  : C:\DX\DATA\98102501.D11
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 11              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	358	1914	2	
2	1.01	fluoride	409.428	308	1373	2	2.36
3	1.10		0.000	419	3418	2	
4	1.46	chloride	4244.636	2112	10297	1	-2.01
5	1.75	nitrite	60701.653	18959	101823	1	-2.42
6	2.58	bromide	532.818	24	106	1	1.57
7	2.82	nitrate	207987.768	44249	370429	1	-3.42
8	3.70		0.000	25	166	1	
9	4.21	phosphate	738.409	28	306	1	0.16
10	4.64		0.000	38	900	1	
11	5.44	sulfate	1684.959	203	2725	1	-0.91
12	7.09	oxalate	331.881	18	383	1	-0.79
13	8.96		0.000	14	385	1	
Totals			276631.552	66755	494225		

File: 98102501.D11 Sample: S98T002961 SAM

```

=====
Sample Name: S98T002961 DUP                      Date: 10/25/1998 12:21:03
Data File  : C:\DX\DATA\98102501.D12
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 12              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	376	2060	2	
2	1.01	fluoride	410.124	319	1375	2	2.36
3	1.09		0.000	423	3471	2	
4	1.46	chloride	4290.286	2131	10411	1	-2.01
5	1.74	nitrite	62436.930	19079	104808	1	-2.79
6	2.57	bromide	513.907	20	83	1	1.31
7	2.81	nitrate	209014.478	44733	372505	1	-3.65
8	3.72		0.000	28	244	1	
9	4.19	phosphate	690.470	30	273	1	-0.32
10	5.44	sulfate	1800.343	214	2943	1	-0.91
Totals			279156.538	67355	498173		

File: 98102501.D12 Sample: S98T002961 DUP

LABCORE Completed Worklist Report for Worklist# 26759

Analyst: vlm

Instrument: IC40S1

Book#: 29N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1	CCB	0	@IC-QC	F	QC	1	<1.20e-2	ug/mL	
1	CCB	0	@IC-QC	CL	QC	1	1.90e-02	0.019 ug/mL	
1	CCB	0	@IC-QC	NO2	QC	1	<1.08e-1	ug/mL	
1	CCB	0	@IC-QC	BR	QC	1	<1.25e-1	ug/mL	
1	CCB	0	@IC-QC	NO3	QC	1	<1.39e-1	ug/mL	
1	CCB	0	@IC-QC	P04	QC	1	<1.20e-1	ug/mL	
1	CCB	0	@IC-QC	S04	QC	1	<1.38e-1	ug/mL	
1	CCB	0	@IC-QC	OxALATE2	QC	1	<1.05e-1	ug/mL	
2	ICV	0	@IC-QC	F	QC	5.90e1	6.29e+01	106.610 % Recovery	
2	ICV	0	@IC-QC	CL	QC	8.00e1	8.07e+01	100.875 % Recovery	
2	ICV	0	@IC-QC	NO2	QC	5.70e2	5.66e+02	99.298 % Recovery	
2	ICV	0	@IC-QC	BR	QC	5.86e2	5.64e+02	96.246 % Recovery	
2	ICV	0	@IC-QC	NO3	QC	5.92e2	5.66e+02	95.608 % Recovery	
2	ICV	0	@IC-QC	P04	QC	5.47e2	5.40e+02	98.720 % Recovery	
2	ICV	0	@IC-QC	S04	QC	6.38e2	6.47e+02	101.411 % Recovery	
2	ICV	0	@IC-QC	OxALATE2	QC	5.44e2	5.67e+02	104.228 % Recovery	
3	CCV	0	@IC-QC	F	QC	6.40e1	6.45e+01	100.781 % Recovery	
3	CCV	0	@IC-QC	CL	QC	9.00e1	9.49e+01	105.444 % Recovery	
3	CCV	0	@IC-QC	NO2	QC	5.53e2	5.82e+02	105.244 % Recovery	
3	CCV	0	@IC-QC	BR	QC	6.30e2	6.23e+02	98.889 % Recovery	
3	CCV	0	@IC-QC	NO3	QC	6.98e2	7.12e+02	102.006 % Recovery	
3	CCV	0	@IC-QC	P04	QC	6.32e2	6.49e+02	102.690 % Recovery	
3	CCV	0	@IC-QC	S04	QC	6.99e2	6.99e+02	100.000 % Recovery	
3	CCV	0	@IC-QC	OxALATE2	QC	5.47e2	5.65e+02	103.291 % Recovery	
4	SAMPLE	S98T002962	0	@IC-01	F-02	LIQUID	N/A	3.207e+02	122.400 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	CL-02	LIQUID	N/A	4.584e+03	173.400 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	NO2-02	LIQUID	N/A	6.323e+04	1102.000 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	BR-02	LIQUID	N/A	< 1.275e+03	1275.000 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	NO3-02	LIQUID	N/A	1.798e+05	1418.000 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	P04-02	LIQUID	N/A	< 1.224e+03	1224.000 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	S04-02	LIQUID	N/A	2.139e+03	1409.000 ug/mL
4	SAMPLE	S98T002962	0	@IC-01	OxALATE2	LIQUID	N/A	< 1.071e+03	1071.000 ug/mL
5	DUP	S98T002962	0	@IC-01	F-02	LIQUID	3.21e+02	3.14e+02	2.205 RPD
5	DUP	S98T002962	0	@IC-01	CL-02	LIQUID	4.58e+03	4.67e+03	1.946 RPD
5	DUP	S98T002962	0	@IC-01	NO2-02	LIQUID	6.32e+04	6.46e+04	2.191 RPD
5	DUP	S98T002962	0	@IC-01	BR-02	LIQUID	<1.28e3	<1.28e3	RPD
5	DUP	S98T002962	0	@IC-01	NO3-02	LIQUID	1.80e+05	1.82e+05	1.105 RPD
5	DUP	S98T002962	0	@IC-01	P04-02	LIQUID	<1.22e3	<1.22e3	RPD
5	DUP	S98T002962	0	@IC-01	S04-02	LIQUID	2.14e+03	2.28e+03	6.335 RPD
5	DUP	S98T002962	0	@IC-01	OxALATE2	LIQUID	<1.07e3	<1.07e3	RPD
6	SPK	S98T002962	0	@IC-01	F-02	LIQUID	5.90e1	5.50e+01	93.220 % Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26759

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SPK	S98T002962	0	QIC-01	CL-02	LIQUID	8.00e1	8.22e+01	102.750 % Recovery
6 SPK	S98T002962	0	QIC-01	NO2-02	LIQUID	5.70e2	6.36e+02	111.579 % Recovery
6 SPK	S98T002962	0	QIC-01	BR-02	LIQUID	5.86e2	5.59e+02	95.392 % Recovery
6 SPK	S98T002962	0	QIC-01	NO3-02	LIQUID	5.92e2	6.33e+02	106.926 % Recovery
6 SPK	S98T002962	0	QIC-01	PO4-02	LIQUID	5.47e2	5.58e+02	102.011 % Recovery
6 SPK	S98T002962	0	QIC-01	SO4-02	LIQUID	6.38e2	6.47e+02	101.411 % Recovery
6 SPK	S98T002962	0	QIC-01	OXALATE2	LIQUID	5.44e2	5.63e+02	103.493 % Recovery
7 SAMPLE	S98T002963	0	QIC-01	F-02	LIQUID	N/A	1.224e+02	122.400 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	CL-02	LIQUID	N/A	5.300e+03	179.400 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	NO2-02	LIQUID	N/A	7.308e+04	1102.000 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	BR-02	LIQUID	N/A	1.275e+03	1275.000 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	NO3-02	LIQUID	N/A	1.526e+05	1418.000 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	PO4-02	LIQUID	N/A	1.224e+03	1224.000 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	SO4-02	LIQUID	N/A	1.921e+03	1408.000 ug/mL
7 SAMPLE	S98T002963	0	QIC-01	OXALATE2	LIQUID	N/A	1.071e+03	1071.000 ug/mL
8 DUP	S98T002963	0	QIC-01	F-02	LIQUID	<1.22e2	<1.22e2	RPD
8 DUP	S98T002963	0	QIC-01	CL-02	LIQUID	5.30e+03	5.12e+03	3.455 RPD
8 DUP	S98T002963	0	QIC-01	NO2-02	LIQUID	7.31e+04	7.09e+04	3.056 RPD
8 DUP	S98T002963	0	QIC-01	BR-02	LIQUID	<1.28e3	<1.28e3	RPD
8 DUP	S98T002963	0	QIC-01	NO3-02	LIQUID	1.53e+05	1.50e+05	1.980 RPD
8 DUP	S98T002963	0	QIC-01	PO4-02	LIQUID	<1.22e3	<1.22e3	RPD
8 DUP	S98T002963	0	QIC-01	SO4-02	LIQUID	1.92e+03	2.66e+03	32.314 RPD
8 DUP	S98T002963	0	QIC-01	OXALATE2	LIQUID	<1.07e3	<1.07e3	RPD
9 SPK	S98T002963	0	QIC-01	F-02	LIQUID	5.90e1	6.53e+01	110.678 % Recovery
9 SPK	S98T002963	0	QIC-01	CL-02	LIQUID	8.00e1	7.97e+01	99.625 % Recovery
9 SPK	S98T002963	0	QIC-01	NO2-02	LIQUID	5.70e2	6.10e+02	107.018 % Recovery
9 SPK	S98T002963	0	QIC-01	BR-02	LIQUID	5.86e2	5.56e+02	94.881 % Recovery
9 SPK	S98T002963	0	QIC-01	NO3-02	LIQUID	5.92e2	6.01e+02	101.520 % Recovery
9 SPK	S98T002963	0	QIC-01	PO4-02	LIQUID	5.47e2	5.45e+02	99.634 % Recovery
9 SPK	S98T002963	0	QIC-01	SO4-02	LIQUID	6.38e2	6.44e+02	100.940 % Recovery
9 SPK	S98T002963	0	QIC-01	OXALATE2	LIQUID	5.44e2	5.50e+02	101.103 % Recovery
11 DUP	S98T002964	0	QIC-01	F-02	LIQUID	3.78e+02	2.66e+02	34.783 RPD
11 DUP	S98T002964	0	QIC-01	CL-02	LIQUID	6.06e+03	4.61e+03	27.179 RPD
11 DUP	S98T002964	0	QIC-01	NO2-02	LIQUID	7.95e+04	6.34e+04	22.533 RPD
11 DUP	S98T002964	0	QIC-01	BR-02	LIQUID	<1.28e3	<1.28e3	RPD
11 DUP	S98T002964	0	QIC-01	NO3-02	LIQUID	1.90e+05	1.50e+05	23.529 RPD
11 DUP	S98T002964	0	QIC-01	PO4-02	LIQUID	<1.22e3	<1.22e3	RPD
11 DUP	S98T002964	0	QIC-01	SO4-02	LIQUID	2.27e+03	1.95e+03	15.166 RPD
11 DUP	S98T002964	0	QIC-01	OXALATE2	LIQUID	<1.07e3	<1.07e3	RPD

Final page for worklist# 26759

Analyst Signature Date

Analyst Signature Date

James M. Lyle 11/2/98
Reviewer Signature Date

Sample# 2964 was reanalyzed on wlist# 26988. *JS* 3/23/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26759

Analyst: VLM Instrument: IC 4051 Book# 29821-A-ICV
28121-A CCV

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 ICV			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 SAMPLE	S98T002962 0		@IC-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
5 DUP	S98T002962 0		@IC-01	LIQUID		
6 SPK	S98T002962 0		@IC-01	LIQUID		
7 SAMPLE	S98T002963 0		@IC-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
8 DUP	S98T002963 0		@IC-01	LIQUID		
9 SPK	S98T002963 0		@IC-01	LIQUID		
10 SAMPLE	S98T002964 0		@IC-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
11 DUP	S98T002964 0		@IC-01	LIQUID		

Final page for worklist # 26759

Valerie Massie 10-27-98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

uploaded 10-29-98
JH Wall
2675900T.CSV

S98T002964 was
reported for poor RPD'S
the rest were validated
11/2/98 Jm Jye

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-A ICV Date: 10/27/1998 10:27:56
 Data File : C:\DX\DATA\98102701.D02
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 2 Detector: CDM-1
 Analyst : *Valerie Mason* Column: AG4A/AS4A anion column
10-27-98

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	64	234	1	
2	1.00	fluoride	106.58	62.883	2147	10822	1 1.01
3	1.48	chloride	100.84	80.673	2277	9973	1 -0.67
4	1.75	nitrite	99.31	566.046	9369	47643	1 -2.05
5	2.60	bromide	96.17	563.534	5896	34118	1 2.36
6	2.93	nitrate	95.66	566.405	6681	46018	1 0.46
7	4.17	phosphate	93.71	539.943	1564	19263	1 -0.63
8	5.44	sulfate	101.44	647.214	5276	63110	1 -0.91
9	7.04	oxalate	104.16	566.627	2271	41421	1 -1.54
Totals			3593.324	35546	272602		

File: 98102701.D02 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1150 TO 1158.

Sample Name: 28N21-A CCV Date: 10/27/1998 10:51:54
 Data File : C:\DX\DATA\98102701.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3 Detector: CDM-1
 Analyst : *Valerie Massie* Column: AG4A/AS4A anion column
10-27-98

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	83	341	2	
2	1.01	fluoride	100.71	64.456	11096	2	1.68
3	1.49	chloride	105.44	94.896	11783	1	-0.22
4	1.77	nitrite	105.21	581.799	9222	1	-1.30
5	2.61	bromide	98.89	623.016	6404	1	2.62
6	2.94	nitrate	102.07	712.418	8256	1	0.68
7	4.20	phosphate	102.47	648.885	1851	1	0.00
8	5.44	sulfate	100.19	699.300	5325	1	-0.91
9	7.09	oxalate	103.29	565.015	2541	1	-0.79
Totals			3989.785	38399	301242		

File: 98102701.D03 Sample: 28N21-A CCV

Sample Name: CCB

Date: 10/27/1998 09:38:27

Data File : C:\DX\DATA\98102701.D01

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 1

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.40	chloride	0.019	17	129	1	-6.04
Totals			0.019	17	129		

File: 98102701.D01 Sample: CCB

Sample Name: S98T002962

Date: 10/27/1998 11:05:51

Data File : C:\DX\DATA\98102701.D04

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 4

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	10201	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	189	1016	2	
2	1.01	fluoride	320.654	128	548	2	2.36
3	1.10		0.000	211	1593	2	
4	1.47	chloride	4583.760	1217	5526	1	-1.57
5	1.75	nitrite	63229.523	10147	52803	1	-2.42
6	2.87	nitrate	179840.591	20204	150840	1	-1.60
7	5.44	sulfate	2138.767	113	1578	1	-0.91
Totals			250113.294	32208	213904		

File: 98102701.D04 Sample: S98T002962

Sample Name: S98T002962 DUP Date: 10/27/1998 11:38:21
 Data File : C:\DX\DATA\98102701.D05
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 5 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	10201	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	173	910	2	
2	1.02	fluoride	314.391	125	538	2	3.03
3	1.10		0.000	203	1543	2	
4	1.47	chloride	4667.348	1221	5630	1	-1.57
5	1.75	nitrite	64634.583	10388	54001	1	-2.42
6	2.87	nitrate	181664.512	20765	152451	1	-1.60
7	5.44	sulfate	2279.742	112	1713	1	-0.91
8	8.96		0.000	14	450	1	
Totals			253560.575	32999	217234		

File: 98102701.D05 Sample: S98T002962 DUP

```

=====
Sample Name: S98T002962 SPIKE                      Date: 10/27/1998 11:52:01
Data File  : C:\DX\DATA\98102701.D06
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 6                 Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1       10201    3000 5Hz   0.00 10.00       30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	203	822	2	
2	1.00	fluoride	5881.006	1965	10011	2	1.01
3	1.46	chloride	12890.867	3400	15987	1	-2.01
4	1.74	nitrite	127505.417	20130	108162	1	-2.79
5	2.56	bromide	56485.209	6090	33851	1	0.79
6	2.84	nitrate	243828.485	27436	208316	1	-2.74
7	3.70		0.000	17	126	1	
8	4.17	phosphate	56364.274	1598	19929	1	-0.63
9	5.39	sulfate	67484.725	4488	65206	1	-1.88
10	7.04	oxalate	56830.103	2474	41129	1	-1.54
Totals			627270.086	67801	503538		

File: 98102701.D06 Sample: S98T002962 SPIKE

```

=====
Sample Name: S98T002963                      Date: 10/27/1998 14:13:27
Data File  : C:\DX\DATA\98102711.D07
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 7           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 10201 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	209	1192	2	
2	1.09		0.000	208	2075	2	
3	1.46	chloride	5300.500	1374	6416	1	-2.01
4	1.73	nitrite	73080.162	11708	61212	1	-3.17
5	2.85	nitrate	152621.771	17744	126999	1	-2.28
6	5.44	sulfate	1920.857	100	1369	1	-0.91
Totals			232923.291	31344	199263		

File: 98102711.D07 Sample: S98T002963

Sample Name: S98T002963 DUP

Date: 10/27/1998 14:27:36

Data File : C:\DX\DATA\98102711.D08

Method: C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 8

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	10201	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	202	1109	2	
2	1.09		0.000	207	2041	2	
3	1.45	chloride	5121.215	1339	6193	1	-2.91
4	1.72	nitrite	70936.525	11389	59380	1	-3.91
5	2.84	nitrate	150326.842	17500	125005	1	-2.74
6	5.44	sulfate	2657.365	121	2074	1	-0.91
Totals			229041.947	30757	195802		

File: 98102711.D08 Sample: S98T002963 DUP

```

=====
Sample Name: S98T002963 SPIKE                      Date: 10/27/1998 14:41:07
Data File  : C:\DX\DATA\98102711.D09
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 9                 Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 10201 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	243	1140	2	
2	1.00	fluoride	6592.760	1860	11239	2	1.01
3	1.45	chloride	13350.494	3531	16575	1	-2.46
4	1.73	nitrite	134671.378	21614	114406	1	-3.17
5	2.55	bromide	56140.630	6097	33639	1	0.26
6	2.83	nitrate	213295.937	24473	180632	1	-3.20
7	4.18	phosphate	55019.819	1575	19440	1	-0.48
8	5.44	sulfate	66964.044	5407	64690	1	-0.91
9	7.04	oxalate	55512.307	2384	40163	1	-1.54
Totals			601547.368	67184	481925		

File: 98102711.D09 Sample: S98T002963 SPIKE

LABCORE Completed Worklist Report for Worklist# 26760

Analyst: kjt

Instrument: IC45S2

Book#: 34N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0		@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 ICV	0		@IC4G-QC F*4	QC	5.89e1	6.01e+01	102.037 % Recovery	
2 ICV	0		@IC4G-QC ACETATE2	QC	1.24e2	1.12e+02	90.323 % Recovery	
2 ICV	0		@IC4G-QC FORMATE2	QC	1.19e2	1.08e+02	90.756 % Recovery	
2 ICV	0		@IC4G-QC GLYCOLT1	QC	1.07e2	1.00e+02	93.458 % Recovery	
3 CCV	0		@IC4G-QC F*4	QC	6.21e1	5.85e+01	94.203 % Recovery	
3 CCV	0		@IC4G-QC ACETATE2	QC	1.23e2	1.15e+02	93.496 % Recovery	
3 CCV	0		@IC4G-QC FORMATE2	QC	1.30e2	1.21e+02	93.077 % Recovery	
3 CCV	0		@IC4G-QC GLYCOLT1	QC	1.02e2	9.74e+01	95.490 % Recovery	
4 SAMPLE	S98T002391	0	@IC4G-01 F*4-01	LIQUID	N/A	< 3.100e-02	3.10e-002	ug/mL
4 SAMPLE	S98T002391	0	@IC4G-01 ACETATE2	LIQUID	N/A	< 5.800e-02	5.80e-002	ug/mL
4 SAMPLE	S98T002391	0	@IC4G-01 FORMATE2	LIQUID	N/A	< 6.200e-02	6.20e-002	ug/mL
4 SAMPLE	S98T002391	0	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 5.600e-02	5.60e-002	ug/mL
5 DUP	S98T002391	0	@IC4G-01 F*4-01	LIQUID	<3.10e-2	<3.10e-2		RPD
5 DUP	S98T002391	0	@IC4G-01 ACETATE2	LIQUID	<5.80e-2	<5.80e-2		RPD
5 DUP	S98T002391	0	@IC4G-01 FORMATE2	LIQUID	<6.20e-2	<6.20e-2		RPD
5 DUP	S98T002391	0	@IC4G-01 GLYCOLT1	LIQUID	<5.60e-2	<5.60e-2		RPD

Final page for worklist# 26760

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 1/26/99

LABCORE Data Entry Template for Worklist# 26760

Analyst: KIT Instrument: IC 4552 Book# 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	ICV				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	SAMPLE	S98T002391 0			@IC4G-01 LIQUID		98000436	BY-105 (2)
					Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1			
5	DUP	S98T002391 0			@IC4G-01 LIQUID			

Final page for worklist # 26760

See attached
_____
Analyst Signature   Date

_____
Analyst Signature   Date

Data Entry Comments:

uploaded 1-25-99   validated 1/25/99 JM Suze
John Worell
26760JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26760

Analyst: KSP Instrument: IC _____ Book# 30020C

Method: LA-533-1105 Rev/Mod F-0

(has been
Rad removed)

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC QC			
2 ICV			@IC4G-QC QC			
3 CCV			@IC4G-QC QC			
4 SAMPLE	S98T002392 0		@IC4G-01 LIQUID		98000436	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
5 DUP	S98T002392 0		@IC4G-01 LIQUID			
6 SAMPLE	S98T002814 0		@IC4G-01 LIQUID		98000437	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
7 DUP	S98T002814 0		@IC4G-01 LIQUID			
8 SAMPLE	S98T002815 0		@IC4G-01 LIQUID		98000437	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
9 DUP	S98T002815 0		@IC4G-01 LIQUID			

} Ran out
of
time

Final page for worklist # 26760

KSPromake 1-12-98
Analyst Signature Date

Analyst Signature Date

S98T002392 in Lab: Blank not FB changing
and running FB P S98T002391

↓
on
new
WKL

27895

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 01/12/1999 13:32:59
 Data File : C:\DX\DATA\99011021.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst : *R. S. Hernandez* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	58068	2688940	1	
2	2.10		0.000	50795	1376878	1	
3	11.36		0.000	7626875	299168525	2	
4	11.73		0.000	8421082	73350602	2	
5	12.03		0.000	13816673	989200637	2	
Totals			0.000	29973493	1365785582		

File: 99011021.D01 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1162 TO 1167

Sample Name: ICV 35N21C Date: 01/12/1999 13:52:51
 Data File : C:\DX\DATA\99011021.D02
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2
 Analyst : *KS Hemdele* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	13301	463200	1	
2	2.24		0.000	3335	71929	2	
3	2.55		0.000	8465	78233	2	
5	2.93		0.000	1901	35874	2	
9	3.93		0.000	2104	39984	2	
13	4.84		0.000	2362	38199	2	
14	5.12		0.000	2839	22438	2	
17	5.55		0.000	2754	40069	2	
18	5.81		0.000	1995	39562	2	
19	6.13		0.000	2601	18583	2	
20	6.43		0.000	2815	49294	2	
21	7.31	fluoride	102.190	1674826	25802418	2	0.00
22	8.03	glycolate	93.590	480834	8908521	2	0.00
23	8.53	acetate	90.090	556125	11180182	2	0.00
24	9.25	formate	90.590	942433	18076111	2	0.00
25	11.33		0.000	7703722	256580347	2	
26	11.71		0.000	6889274	84146398	2	
27	12.05		0.000	8756715	702221172	3	
28	15.25		0.000	27411	1992279	4	

Totals 379.598 27075810 1109804793

File: 99011021.D02 Sample: ICV 35N21C

Data Reprocessed On 01/12/1999 14:21:22

Sample Name: CCV 34N21A Date: 01/12/1999 14:13:20
 Data File : C:\DX\DATA\99011021.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
 Analyst : *K.S. Thorndike* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
19	4.33		0.000	2072	33458	2	
21	4.80		0.000	2193	19570	2	
22	5.09		0.000	2255	17970	2	
23	5.29		0.000	2620	21588	2	
24	5.44		0.000	3151	20219	2	
25	5.60		0.000	3396	33504	2	
28	6.12		0.000	1399	18650	2	
30	7.28	fluoride	58.505	1647884	25076066	2	0.00
31	8.00	glycolate	97.380	465304	8668352	2	0.00
32	8.51	acetate	115.148	585262	11524453	2	0.00
33	9.23	formate	120.924	1061882	20327563	2	0.00
34	11.33		0.000	7723166	265441435	2	
35	11.73		0.000	6651651	72217492	2	
36	12.03		0.000	8203741	711980290	2	

Totals	391.957	26355977	1115400609

Data Reprocessed On 01/25/1999 15:02:54

```

=====
Sample Name: S98T002391          Date: 01/12/1999 14:53:33
Data File  : E:\DATA\99011021.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====
  
```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
12	3.16		0.000	2620	21908	2	
13	3.33		0.000	3548	20173	2	
31	6.85		0.000	1780	36330	2	
32	7.25	fluoride	0.010	4322	84816	2	0.00
33	7.71		0.000	2141	39674	2	
35	8.48	acetate	-0.009	4846	134108	2	0.00
36	9.25	formate	0.013	7371	163209	2	0.00
38	11.33		0.000	7588944	240728759	2	
39	11.73		0.000	6370977	77670303	2	
40	12.03		0.000	7932735	696758127	3	
41	17.55		0.000	2725	106241	4	

Totals 0.014 21922009 1015763648

Sample Name: S98T002391 DUP Date: 01/12/1999 15:13:49
 Data File : C:\DX\DATA\99011021.D06
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 6 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	9120	273832	1	
7	2.97		0.000	3360	44561	2	
8	3.13		0.000	3294	29304	2	
11	3.51		0.000	2886	16148	2	
12	3.65		0.000	2819	19745	2	
13	3.76		0.000	3208	20476	2	
16	4.37		0.000	3386	17687	2	
17	4.51		0.000	4033	37107	2	
19	4.81		0.000	3708	44807	2	
20	5.21		0.000	2227	26444	2	
29	6.93		0.000	1975	16943	2	
30	7.31	fluoride	0.010	4645	68088	2	0.00
31	7.73		0.000	2022	27082	2	
32	8.16	glycolate	-0.017	1216	19300	2	0.00
33	8.59	acetate	-0.013	4713	91394	2	0.00
34	9.28	formate	0.010	5537	114683	2	0.00
37	11.07		0.000	3552080	94784338	2	
38	11.36		0.000	7630458	136507365	2	
39	11.73		0.000	6450211	70056821	2	
40	12.03		0.000	8194589	704529915	3	
41	17.92		0.000	3015	79765	4	
Totals			-0.011	25888502	1006825805		

LABCORE Completed Worklist Report for Worklist# 26761

Analyst: adp

Instrument: IC45S2

Book#: 34N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: B1105

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2	ICV	0	@IC4G-QC F*4	QC	5.89e1	6.14e+01	104.244 % Recovery	
2	ICV	0	@IC4G-QC ACETATE2	QC	1.24e2	1.16e+02	93.548 % Recovery	
2	ICV	0	@IC4G-QC FORMATE2	QC	1.19e2	1.13e+02	94.958 % Recovery	
2	ICV	0	@IC4G-QC GLYCOLTI	QC	1.07e2	1.05e+02	98.131 % Recovery	
3	CCV	0	@IC4G-QC F*4	QC	6.21e1	5.85e+01	94.203 % Recovery	
3	CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.138e02	92.520 % Recovery	
3	CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.200e02	92.308 % Recovery	
3	CCV	0	@IC4G-QC GLYCOLTI	QC	1.02e2	9.695e01	95.049 % Recovery	
4	SAMPLE	S98T002816	@IC4G-01 F*4-01	LIQUID	N/A	2.508e+02	159.700	ug/mL
4	SAMPLE	S98T002816	@IC4G-01 ACETATE2	LIQUID	N/A	1.375e+03	298.800	ug/mL
4	SAMPLE	S98T002816	@IC4G-01 FORMATE2	LIQUID	N/A	1.040e+03	319.400	ug/mL
4	SAMPLE	S98T002816	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 2.885e+02	288.500	ug/mL
5	DUP	S98T002816	@IC4G-01 F*4-01	LIQUID	2.51e+02	2.62e+02	4.288	RPD
5	DUP	S98T002816	@IC4G-01 ACETATE2	LIQUID	1.37e+03	1.31e+03	4.478	RPD
5	DUP	S98T002816	@IC4G-01 FORMATE2	LIQUID	1.04e+03	1.06e+03	1.905	RPD
5	DUP	S98T002816	@IC4G-01 GLYCOLTI	LIQUID	<2.88e+02	<2.88e2		RPD
6	SAMPLE	S98T002817	@IC4G-01 F*4-01	LIQUID	N/A	3.843e+02	159.700	ug/mL
6	SAMPLE	S98T002817	@IC4G-01 ACETATE2	LIQUID	N/A	1.558e+03	298.800	ug/mL
6	SAMPLE	S98T002817	@IC4G-01 FORMATE2	LIQUID	N/A	1.130e+03	319.400	ug/mL
6	SAMPLE	S98T002817	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 2.885e+02	288.500	ug/mL
7	DUP	S98T002817	@IC4G-01 F*4-01	LIQUID	3.84e+02	3.88e+02	1.036	RPD
7	DUP	S98T002817	@IC4G-01 ACETATE2	LIQUID	1.56e+03	1.59e+03	1.905	RPD
7	DUP	S98T002817	@IC4G-01 FORMATE2	LIQUID	1.13e+03	1.16e+03	2.620	RPD
7	DUP	S98T002817	@IC4G-01 GLYCOLTI	LIQUID	<2.88e2	<2.88e2		RPD
8	SAMPLE	S98T002818	@IC4G-01 F*4-01	LIQUID	N/A	1.416e+02	65.750	ug/mL
8	SAMPLE	S98T002818	@IC4G-01 ACETATE2	LIQUID	N/A	1.525e+03	123.000	ug/mL
8	SAMPLE	S98T002818	@IC4G-01 FORMATE2	LIQUID	N/A	1.073e+03	131.500	ug/mL
8	SAMPLE	S98T002818	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 1.188e+02	118.800	ug/mL
9	DUP	S98T002818	@IC4G-01 F*4-01	LIQUID	1.41e+02	1.41e+02	0.000	RPD
9	DUP	S98T002818	@IC4G-01 ACETATE2	LIQUID	1.53e+03	1.53e+03	0.000	RPD
9	DUP	S98T002818	@IC4G-01 FORMATE2	LIQUID	1.07e+03	1.08e+03	0.930	RPD
9	DUP	S98T002818	@IC4G-01 GLYCOLTI	LIQUID	<1.19e2	<1.19e2		RPD

Final page for worklist# 26761

Units shown for QC (BLK/BKG) may not reflect the actual units.

John M. Lye 1/18/99

LABCORE Data Entry Template for Worklist# 26761

Analyst: ADP Instrument: IC Book# 3SN21-C LCS
 Method: LA-533-1105 Rev/Mod F-0 34N21-A CCV
 Worklist Comment: BY-105 (2) EIC4G-01 (has been Rad screened)

THB 10/23/98

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	ICV				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	SAMPLE	S98T002816 0			@IC4G-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
5	DUP	S98T002816 0			@IC4G-01	LIQUID		
6	SAMPLE	S98T002817 0			@IC4G-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
7	DUP	S98T002817 0			@IC4G-01	LIQUID		
8	SAMPLE	S98T002818 0			@IC4G-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
9	DUP	S98T002818 0			@IC4G-01	LIQUID		

Final page for worklist # 26761

Anthony P. Wenzel 1-15-98
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 1-18-99 Validated 1/18/99 JMW
John Wenzel
26761JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

=====

Sample Name: 35N21-C LCS	Date: 01/15/1999 09:51:35
Data File : C:\DX\DATA\99011501.D04	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System 2 Inject#: 4	Detector: CDM-2
Analyst : <i>Anthony Puro</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 SHz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	15711	480772	1	
4	2.64		0.000	2439	30501	2	
5	2.87		0.000	3361	30099	2	
6	3.08		0.000	4388	25686	2	
7	3.16		0.000	4094	59976	2	
8	3.60		0.000	2963	28491	2	
9	3.73		0.000	3365	16447	2	
11	3.91		0.000	3009	19295	2	
12	4.04		0.000	3431	22474	2	
13	4.15		0.000	3165	35979	2	
14	4.47		0.000	2906	43423	3	
19	5.23		0.000	3436	81129	2	
22	6.23		0.000	1842	17152	2	
25	7.31	fluoride	61.398 ^{104.24}	1779312	26353818	2	0.00
26	8.03	glycolate	104.913 ^{98.05}	513727	9341763	2	0.00
27	8.53	acetate	116.018 ^{73.56}	597800	11609496	2	0.00
28	9.23	formate	112.750 ^{94.15}	988284	18929545	2	0.00
29	11.33		0.000	8012405	273024029	2	
30	11.73		0.000	6607802	82042766	2	
31	12.03		0.000	8518293	657258539	2	
32	18.69		0.000	2400	31280	1	
Totals			395.079	27074133	1079482661		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1170 TO 1186.

1171

```

=====
Sample Name: 43N21-A CCV 34N21-A Date: 01/15/1999 09:30:21
Data File : C:\DX\DATA\99011501.D03
Method : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
Analyst : Column: AG14A-SC, AS14A-SC, SRS
=====
    
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External 1 101 5700 5Hz 0.00 19.00 16000
    
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	14842	224523	2	
2	0.60		0.000	20070	252781	2	
3	1.53		0.000	1962	23448	1	
4	2.10		0.000	19208	324205	2	
5	2.27		0.000	14472	146363	2	
6	2.73		0.000	3329	16984	1	
7	2.93		0.000	3314	29804	1	
9	3.59		0.000	4039	60600	2	
10	3.91		0.000	2682	46502	2	
18	5.68		0.000	3675	29632	2	
19	5.88		0.000	2430	34228	2	
21	7.31	fluoride	58.496	1738459	25072248	2	0.00
22	8.00	glycolate	96.954	479701	8630428	2	0.00
23	8.51	acetate	113.850	599121	11397499	2	0.00
24	9.20	formate	120.045	1097563	20176999	2	0.00
25	11.33		0.000	8523370	418703174	2	
26	12.03		0.000	9790725	671449233	2	
Totals			389.345	22318963	1156618652		

1172

```

=====
Sample Name: BLANK                      Date: 01/15/1999 08:42:46
Data File  : C:\DX\DATA\99011501.D01
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 1
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700  5Hz    0.00 19.00    16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.33		0.000	502412	18666840	1	
2	3.29		0.000	15506	301800	1	
5	6.83		0.000	17064	471600	1	
6	8.85		0.000	4778	43800	1	
7	11.36		0.000	8020696	315781055	2	
8	11.71		0.000	11431434	127934065	2	
9	12.03		0.000	18737884	1328478191	3	
10	15.63		0.000	45553	4269887	4	
11	17.79		0.000	6879	96124	4	
Totals			0.000	38782205	1796043362		

File: 99011501.D01 Sample: BLANK

1173

Data Reprocessed On 01/18/1999 14:39:01

```

=====
Sample Name: S98T002816 SAM                      Date: 01/15/1999 10:37:08
Data File  : E:\DATA\99011501.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151  5700  5Hz  5.96 10.69      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	10787	200515	2	
2	0.73		0.000	7014	71458	2	
3	2.28		0.000	6500	115710	1	
4	2.85		0.000	7143	106172	2	
5	2.97		0.000	7852	60748	2	
6	3.13		0.000	8236	47873	2	
7	3.23		0.000	9027	103676	2	
8	3.56		0.000	9636	77671	2	
9	3.65		0.000	10307	211705	2	
10	4.07		0.000	8547	139271	2	
11	4.36		0.000	5879	61991	2	
12	4.57		0.000	4159	55331	2	
13	4.84		0.000	3333	41455	3	
17	5.88		0.000	1551	25342	1	
19	6.69		0.000	3457	27349	2	
20	7.28	fluoride	250.822	108467	1758953	2	0.00
21	8.03	glycolate	-30.547	10022	117972	2	0.00
22	8.51	acetate	1375.996	145443	2878777	2	0.00
23	9.20	formate	1040.315	178673	3344487	2	0.00
24	9.97		0.000	27369	437636	1	
25	10.91		0.000	6051935	58902106	2	
26	11.20		0.000	38684203	487520262	2	
27	11.89		0.000	68246837	1344989483	3	
28	17.95		0.000	2464	29786	4	

```

-----
Totals                2636.586  113558842  1901325729
-----

```

Data Reprocessed On 01/18/1999 14:40:04

```

=====
Sample Name: S98T002816  DUP                               Date: 01/15/1999 11:17:07
Data File  : E:\DATA\99011501.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7                        Detector:CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700  5Hz   6.38 10.67      16000
-----
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.63		0.000	11300	110660	1	
3	3.16		0.000	6234	241840	2	
4	3.47		0.000	1806	74614	2	
17	5.77		0.000	1420	37760	1	
18	7.23	fluoride	262.423	108007	1856556	2	0.00
19	7.97	glycolate	6.513	13620	180033	2	0.00
20	8.43	acetate	1308.657	134113	2748920	2	0.00
21	9.12	formate	1057.843	179868	3401794	2	0.00
22	9.87		0.000	35481	673388	2	
23	10.91		0.000	7663124	95870729	2	
24	11.20		0.000	39153110	493629880	2	
25	11.89		0.000	70112895	1285796813	3	
26	18.43		0.000	2710	48363	4	

Totals 2635.436 117423687 1884671349

File: 99011501.D07 Sample: S98T002816 DUP

1176

Data Reprocessed On 01/18/1999 14:40:38

```

=====
Sample Name: S98T002817 SAM                      Date: 01/15/1999 11:38:02
Data File  : E:\DATA\99011501.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8                Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151  5700  5Hz  6.08 10.61      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	11425	123650	3	
5	2.20		0.000	3165	34948	1	
6	2.51		0.000	4405	25054	2	
7	2.68		0.000	5649	35833	2	
8	2.79		0.000	4461	37429	2	
9	2.99		0.000	3838	46643	2	
10	3.24		0.000	4821	19091	2	
20	4.33		0.000	2727	25104	1	
22	4.99		0.000	3164	38758	2	
26	5.68		0.000	5638	50901	2	
27	5.85		0.000	6405	115381	2	
28	7.31	fluoride	384.276	182798	2882466	2	0.00
29	8.53	acetate	1558.026	155362	3229751	2	0.00
30	9.23	formate	1129.697	197499	3636765	2	0.00
31	10.00		0.000	27610	422584	1	
32	10.91		0.000	6288749	59893174	2	
33	11.17		0.000	34254508	480880418	2	
34	11.89		0.000	74555555	1368487905	3	
35	17.60		0.000	6769	47624	4	
36	17.76		0.000	7244	81123	4	
37	18.03		0.000	4856	61790	4	
38	18.24		0.000	4988	73419	4	
39	18.75		0.000	3081	18069	4	

```

-----
Totals          3072.000  115744719 1920267879

```

Data Reprocessed On 01/18/1999 14:45:17

```

=====
Sample Name: S98T002817   SAM                               Date: 01/15/1999 11:38:02
Data File  : E:\DATA\99011501.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700  5Hz   6.67 10.61    16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	11425	123650	3	
5	2.20		0.000	3165	34948	1	
6	2.51		0.000	4405	25054	2	
7	2.68		0.000	5649	35833	2	
8	2.79		0.000	4461	37429	2	
9	2.99		0.000	3838	46643	2	
10	3.24		0.000	4821	19091	2	
20	4.33		0.000	2727	25104	1	
22	4.99		0.000	3164	38758	2	
26	5.68		0.000	5638	50901	2	
27	5.85		0.000	6405	115381	2	
28	7.31	fluoride	384.276	182798	2882466	2	0.00
29	8.53	acetate	1558.026	155362	3229751	2	0.00
30	9.23	formate	1129.697	197499	3636765	2	0.00
31	10.00		0.000	27610	422584	1	
32	10.91		0.000	6288749	59893174	2	
33	11.17		0.000	34254508	480880418	2	
34	11.89		0.000	74555555	1368487905	3	
35	17.60		0.000	6769	47624	4	
36	17.76		0.000	7244	81123	4	
37	18.03		0.000	4856	61790	4	
38	18.24		0.000	4988	73419	4	
39	18.75		0.000	3081	18069	4	
Totals			3072.000	115744719	1920267879		

Data Reprocessed On 01/18/1999 14:45:59

```

=====
Sample Name: S98T002817 DUP                      Date: 01/15/1999 11:57:59
Data File  : E:\DATA\99011501.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      5151   5700  5Hz   6.29 10.63      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.13		0.000	2898	22660	1	
5	2.07		0.000	4280	57375	2	
9	2.57		0.000	4510	38408	2	
10	2.75		0.000	4300	30125	2	
11	2.88		0.000	5411	51544	2	
13	3.20		0.000	4746	32565	2	
14	3.40		0.000	4702	43920	2	
15	3.55		0.000	3507	40520	2	
16	3.85		0.000	3448	26109	2	
17	4.04		0.000	4092	21490	2	
18	4.17		0.000	3063	21579	2	
19	4.44		0.000	3002	20950	1	
22	4.92		0.000	4160	44748	1	
24	5.37		0.000	3301	19741	2	
25	5.49		0.000	3777	43745	2	
26	5.73		0.000	5409	34838	2	
27	5.85		0.000	6125	43314	2	
28	5.97		0.000	6101	80408	2	
29	6.59		0.000	4658	39192	2	
30	7.28	fluoride	388.388	184552	2917105	2	0.00
31	8.51	acetate	1590.663	157142	3292669	2	0.00
32	9.20	formate	1159.219	199405	3733327	2	0.00
33	9.97		0.000	31743	490220	1	
34	10.91		0.000	3957364	12350010	1	
35	11.17		0.000	30968143	277785400	1	
36	11.71		0.000	782400	889000	1	
37	11.89		0.000	68417323	749680400	1	
Totals			3138.269	104779562	1051851359		

Data Reprocessed On 01/18/1999 14:46:26

```

=====
Sample Name: S98T002818 SAM                      Date: 01/15/1999 12:18:21
Data File  : E:\DATA\99011501.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151  5700  5Hz  6.26 10.56      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	9458	201432	1	
4	2.19		0.000	3561	39124	1	
7	2.65		0.000	5273	31940	2	
8	2.81		0.000	8673	61110	2	
9	2.95		0.000	7462	46905	2	
10	3.07		0.000	5447	31994	2	
11	3.21		0.000	7381	46787	2	
12	3.35		0.000	5911	26885	2	
13	3.44		0.000	6785	43962	2	
14	3.56		0.000	5810	40182	2	
15	3.69		0.000	6680	42405	2	
16	3.85		0.000	7559	58299	2	
17	4.00		0.000	6773	38889	2	
18	4.13		0.000	6903	38232	2	
19	4.21		0.000	7293	46213	2	
20	4.35		0.000	7302	53343	2	
21	4.49		0.000	8477	113270	2	
22	4.80		0.000	9431	82068	2	
23	4.95		0.000	8705	43248	2	
24	5.01		0.000	9090	82668	2	
25	5.20		0.000	10639	128882	2	
26	5.45		0.000	10793	163632	2	
27	5.72		0.000	10632	280963	2	
28	6.69		0.000	5528	86182	2	
29	7.31	fluoride	141.609	49621	840639	2	0.00
30	8.08	glycolate	6.740	13567	180412	2	0.00
31	8.53	acetate	1525.821	160718	3167663	2	0.00
32	9.23	formate	1073.474	184927	3452902	2	0.00
33	10.00		0.000	28473	432664	1	
34	10.91		0.000	6372889	59251355	2	
35	11.20		0.000	41111332	485536684	2	
36	11.89		0.000	68206015	1250778032	2	

```

-----
Totals                2747.645  116309110  1805468967
-----

```

Data Reprocessed On 01/18/1999 14:47:12

Sample Name: S98T002818 DUP	Date: 01/15/1999 12:38:55
Data File : E:\DATA\99011501.D11	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 11	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	5700	5Hz	6.29	10.51	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	9936	229018	1	
4	2.00		0.000	5539	27003	2	
5	2.29		0.000	5342	76865	2	
6	2.59		0.000	4300	20893	2	
7	2.68		0.000	5061	51882	2	
8	2.92		0.000	8953	86363	2	
9	3.08		0.000	5341	49525	2	
10	3.32		0.000	3472	59799	2	
11	3.73		0.000	2829	37341	2	
12	4.00		0.000	3436	33097	2	
13	4.31		0.000	2254	18172	2	
18	4.91		0.000	904	17600	2	
19	5.28		0.000	2340	17360	2	
21	5.56		0.000	5911	43831	2	
22	5.71		0.000	6394	76075	2	
23	5.93		0.000	6406	102485	2	
25	6.69		0.000	4636	43743	2	
26	6.91		0.000	5070	50120	2	
27	7.31	fluoride	141.648	51441	840967	2	0.00
28	8.05	glycolate	-18.490	12811	138159	2	0.00
29	8.53	acetate	1533.489	159543	3182445	2	0.00
30	9.23	formate	1077.930	186299	3467473	2	0.00
31	10.00		0.000	26963	423740	1	
32	10.91		0.000	4015520	18126000	1	
33	11.20		0.000	37483457	278706400	1	
34	11.71		0.000	860667	1663400	1	
35	11.89		0.000	62599707	329332790	1	
Totals			2734.577	105484532	636922544		

LABCORE Completed Worklist Report for Worklist# 26762

Analyst: adp

Instrument: IC45S2

Book#: 131N20B

Method: LA-533-105 *QMF 11/2/98* Rev/Mod *F-0*

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 ICV	0	@IC4G-QC F*4	QC	5.80e1	5.78e+01	99.685 % Recovery	
2 ICV	0	@IC4G-QC ACETATE2	QC	1.17e2	1.17e+02	100.000 % Recovery	
2 ICV	0	@IC4G-QC FORMATE2	QC	1.27e2	1.12e+02	88.189 % Recovery	
2 ICV	0	@IC4G-QC GLYCOLT1	QC	1.01e2	9.60e+01	95.050 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	5.90e1	6.09e+01	103.220 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.17e2	1.12e+02	95.726 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.32e2	1.32e+02	100.000 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.16e2	1.24e+02	106.897 % Recovery	
4 SAMPLE	S98T002819	@IC4G-01 F*4-01	LIQUID	N/A	2.768e+01	18.790 ug/mL	
4 SAMPLE	S98T002819	@IC4G-01 ACETATE2	LIQUID	N/A	2.205e+03	35.150 ug/mL	
4 SAMPLE	S98T002819	@IC4G-01 FORMATE2	LIQUID	N/A	1.292e+03	37.570 ug/mL	
4 SAMPLE	S98T002819	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 3.394e+01	33.940 ug/mL	
5 DUP	S98T002819	@IC4G-01 F*4-01	LIQUID	2.77e+01	2.69e+01	2.930 RPD	
5 DUP	S98T002819	@IC4G-01 ACETATE2	LIQUID	2.20e+03	2.35e+03	6.593 RPD	
5 DUP	S98T002819	@IC4G-01 FORMATE2	LIQUID	1.29e+03	1.38e+03	6.742 RPD	
5 DUP	S98T002819	@IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD	
6 SPK	S98T002819	@IC4G-01 F*4-01	LIQUID	5.80e1	2.35e+01	40.517 % Recovery	
6 SPK	S98T002819	@IC4G-01 ACETATE2	LIQUID	1.17e2	1.03e+02	88.034 % Recovery	
6 SPK	S98T002819	@IC4G-01 FORMATE2	LIQUID	1.27e2	1.43e+02	112.598 % Recovery	
6 SPK	S98T002819	@IC4G-01 GLYCOLT1	LIQUID	1.01e2	1.31e+02	129.703 % Recovery	
7 SAMPLE	S98T002820	@IC4G-01 F*4-01	LIQUID	N/A	2.370e+01	18.790 ug/mL	
7 SAMPLE	S98T002820	@IC4G-01 ACETATE2	LIQUID	N/A	2.313e+03	35.150 ug/mL	
7 SAMPLE	S98T002820	@IC4G-01 FORMATE2	LIQUID	N/A	1.310e+03	37.570 ug/mL	
7 SAMPLE	S98T002820	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 3.394e+01	33.940 ug/mL	
8 DUP	S98T002820	@IC4G-01 F*4-01	LIQUID	2.37e+01	2.34e+01	1.274 RPD	
8 DUP	S98T002820	@IC4G-01 ACETATE2	LIQUID	2.31e+03	2.32e+03	0.432 RPD	
8 DUP	S98T002820	@IC4G-01 FORMATE2	LIQUID	1.31e+03	1.28e+03	2.317 RPD	
8 DUP	S98T002820	@IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD	
9 SAMPLE	S98T002821	@IC4G-01 F*4-01	LIQUID	N/A	< 1.879e+01	18.790 ug/mL	
9 SAMPLE	S98T002821	@IC4G-01 ACETATE2	LIQUID	N/A	2.489e+03	35.150 ug/mL	
9 SAMPLE	S98T002821	@IC4G-01 FORMATE2	LIQUID	N/A	9.557e+02	37.570 ug/mL	
9 SAMPLE	S98T002821	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 3.394e+01	33.940 ug/mL	
10 DUP	S98T002821	@IC4G-01 F*4-01	LIQUID	<1.88e1	<1.88e1	RPD	
10 DUP	S98T002821	@IC4G-01 ACETATE2	LIQUID	2.49e+03	2.60e+03	4.322 RPD	
10 DUP	S98T002821	@IC4G-01 FORMATE2	LIQUID	9.56e+02	9.79e+02	2.377 RPD	
10 DUP	S98T002821	@IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26762

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 26762

Analyst Signature   Date

Analyst Signature   Date

Reviewer Signature   Date 11/2/98

LABCORE Data Entry Template for Worklist# 26762

Analyst: ADD Instrument: IC 4552 Book# 131N20-C CCV
Method: LA-533-V105 ^{jmF 11/2/98} Rev/Mod F-0 131N20-B CCV
Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 ICV			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T002819 0		@IC4G-01 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
5 DUP	S98T002819 0		@IC4G-01 LIQUID			
6 SPK	S98T002819 0		@IC4G-01 LIQUID			
7 SAMPLE	S98T002820 0		@IC4G-01 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
8 DUP	S98T002820 0		@IC4G-01 LIQUID			
9 SAMPLE	S98T002821 0		@IC4G-01 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
10 DUP	S98T002821 0		@IC4G-01 LIQUID			

Final page for worklist # 26762

Anthony Perrotta 1029-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 11-2-98 validated 11/2/98 jmLuge
JH Uebell
267620CT.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Data Reprocessed On 10/29/1998 23:10:19

Sample Name: 131N20-B ICV Date: 10/29/1998 23:00:02
 Data File : C:\DX\DATA\98102911.D04
 Method : C:\DX\METHOD\GLYCOLIC.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-1
 Analyst : Ed Colvin Column: AG4A-SC, AS4A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 4500 5Hz 0.00 15.00 0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.10		0.000	44	295	1	
2	2.43	fluoride	57.782	99.62	14874	2	0.00
3	2.75	acetate	116.740	99.78	5693	2	0.00
4	2.93	glycolate	96.004	95.05	5163	2	0.00
5	3.61	formate	111.593	87.37	10616	1	0.00
6	6.40		0.000	2865	173187	1	
7	8.09		0.000	572	7596	2	
8	8.27		0.000	716	6750	2	
9	8.43		0.000	795	53217	2	
Totals			382.119	9862	277391		

File: 98102911.D04 Sample: 131N20-B ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1197 TO 1199.

1190
 LAD
 3/25/99

Data Reprocessed On 10/30/1998 00:04:51

```

=====
Sample Name: 130N20-C CCV                      Date: 10/20/1998 23:51:05
Data File  : C:\DX\DATA\98102911.D06
Method     : C:\DX\METHOD\GLYCOLIC.MBT
ACI Address: 1 System: 2 Inject#: 6
Analyst    : Ed Colvin      Column: AG4A-SC,AS4A-SC, SRS
Detector: CDM-1
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 4500 5Hz 0.00 15.00 0

***** Peak Report: All Peaks. *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.10		0.000	49	316	1	
2	2.41	fluoride	60.890	103.20	15716	2	0.00
3	2.73	acetate	112.338	96.02	5476	2	0.00
4	2.90	glycolate	123.786	106.71	6761	2	0.00
5	3.56	formate	131.766	99.82	12666	1	0.00
6	6.40		0.000	3586	281460	2	
7	7.95		0.000	927	35187	2	
8	8.25		0.000	297	2195	2	
Totals			428.779	10469	359777		

File: 98102911.D06 Sample: 130N20-C CCV

Data Reprocessed On 11/02/1998 14:05:47

```
=====
Sample Name: BLANK                      Date: 10/29/1998 22:43:28
Data File  : E:\DATA\98102911.D03
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 3      Detector: CDM-1
Analyst    : Ed Colvin                  Column: AG4A-SC, AS4A-SC, SRS
=====
```

```
=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1  4500  5Hz  0.00 15.00          0
=====
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	35	459	1	
2	1.38		0.000	126	4765	1	
3	6.02		0.000	8265	560364	2	
4	8.28		0.000	1716	17265	2	
5	8.47		0.000	1857	112093	2	

```
-----
Totals          0.000          11999          694946
=====
```

File: 98102911.D03 Sample: BLANK

Data Reprocessed On 11/02/1998 13:48:55

Sample Name: S98T002819 SAM	Date: 10/30/1998 00:25:35
Data File : E:\DATA\98102911.D07	
Method : C:\DX\METHOD\GLYCOLIC.MET	
ACI Address: 1 System: 2 Inject#: 7	Detector: CDM-1
Analyst : Ed Colvin	Column: AG4A-SC, AS4A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	606	4500	5Hz	0.00	15.00		0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.96		0.000	103	834	1	
2	1.61		0.000	43	1643	1	
3	2.39	fluoride	27.680	108	625	2	0.00
4	2.67	acetate	2204.869	1789	17603	2	0.00
5	3.47	formate	1291.990	2566	21237	1	0.00
6	4.37		0.000	497	4862	1	
7	5.67		0.000	34759	228704	2	
8	6.01		0.000	83535	1398816	2	
9	7.51		0.000	96047	3472048	2	
Totals			3524.539	219447	5146372		

File: 98102911.D07 Sample: S98T002819 SAM

```

=====
Sample Name: S98T002819 DUP                      Date: 10/30/1998 00:43:39
Data File  : E:\DATA\98102911.D08
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 8                Detector:CDM-1
Analyst    : Ed Colvin                            Column: AG4A-SC,AS4A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           606  4500  5Hz  0.00 15.00          0
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	75	520	1	
2	1.68		0.000	37	1388	1	
3	2.41	fluoride	26.869	101	589	2	0.00
4	2.71	acetate	2351.834	1902	18722	2	0.00
5	3.54	formate	1380.376	2685	22761	1	0.00
6	4.46		0.000	514	5182	1	
7	5.70		0.000	35386	204433	2	
8	6.01		0.000	84060	1413015	2	
9	7.51		0.000	95751	3556129	2	
Totals			3759.079	220512	5222738		

File: 98102911.D08 Sample: S98T002819 DUP

Data Reprocessed On 11/02/1998 13:52:16

```

=====
Sample Name: S98T002819 SPK                      Date: 10/30/1998 00:59:20
Data File  : E:\DATA\98102911.D09
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 9              Detector:CDM-1
Analyst    : Ed Colvin                          Column: AG4A-SC,AS4A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606   4500  5Hz   0.00 15.00          0
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	77	577	1	
2	1.61		0.000	30	1161	1	
3	2.41	fluoride	145.424	945	5850	2	0.00
4	2.71	acetate	2719.120	2439	21486	2	0.00
5	2.89	glycolate	656.877	839	5936	2	0.00
6	3.55	formate	2007.783	3994	33699	1	0.00
7	4.48		0.000	507	5030	1	
8	5.70		0.000	35539	211790	2	
9	6.01		0.000	92472	1393456	2	
10	7.50		0.000	95760	3546279	2	
Totals			5529.203	232602	5225265		

File: 98102911.D09 Sample: S98T002819 SPK

Data Reprocessed On 11/02/1998 13:53:10

Sample Name: S98T002820 SAM	Date: 10/30/1998 01:23:02
Data File : E:\DATA\98102911.D10	
Method : C:\DX\METHOD\GLYCOLIC.MET	
ACI Address: 1 System: 2 Inject#: 10	Detector: CDM-1
Analyst : Ed Colvin	Column: AG4A-SC, AS4A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	606	4500	5Hz	0.00	15.00		0

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	76	571	1	
2	1.73		0.000	39	1450	1	
3	2.41	fluoride	23.699	81	449	2	0.00
4	2.69	acetate	2313.278	1912	18429	2	0.00
5	3.51	formate	1310.130	2580	21550	1	0.00
6	4.40		0.000	491	4883	1	
7	5.67		0.000	33489	193008	2	
8	6.01		0.000	82123	1240691	2	
9	7.55		0.000	89128	3089775	1	
Totals			3647.108	209919	4570805		

File: 98102911.D10 Sample: S98T002820 SAM

```

=====
Sample Name: S98T002820 DUP                      Date: 10/30/1998 01:39:44
Data File  : E:\DATA\98102911.D11
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 11              Detector: CDM-1
Analyst    : Ed Colvin                           Column: AG4A-SC,AS4A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           606 4500 5Hz 0.00 15.00          0
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	71	497	1	
2	1.75		0.000	31	1201	1	
3	2.41	fluoride	23.411	79	436	2	0.00
4	2.71	acetate	2318.614	1947	18469	2	0.00
5	3.55	formate	1283.679	2565	21094	1	0.00
6	4.47		0.000	478	4786	1	
7	5.70		0.000	34400	206236	2	
8	6.01		0.000	84073	1402414	2	
9	7.55		0.000	90940	3357646	2	
Totals			3625.704	214583	5012780		

File: 98102911.D11 Sample: S98T002820 DUP

Data Reprocessed On 11/02/1998 13:55:51

```

=====
Sample Name: S98T002821 SAM                      Date: 10/30/1998 01:56:16
Data File  : E:\DATA\98102911.D12
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 12              Detector: CDM-1
Analyst    : Ed Colvin                          Column: AG4A-SC,AS4A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1          606    4500  5Hz   0.00 15.00          0
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	63	446	1	
2	1.80		0.000	27	1060	1	
3	2.43	fluoride	15.673	25	94	1	0.00
4	2.71	acetate	2489.200	2072	19761	1	0.00
5	3.55	formate	955.712	1832	15475	1	0.00
6	4.48		0.000	352	3467	1	
7	5.70		0.000	32007	194806	2	
8	6.01		0.000	84761	1399275	2	
9	7.58		0.000	86869	3122661	2	
Totals			3460.585	208008	4757045		

File: 98102911.D12 Sample: S98T002821 SAM

Data Reprocessed On 11/02/1998 13:56:48

```

=====
Sample Name: S98T002821 DUP                      Date: 10/30/1998 02:12:20
Data File  : E:\DATA\98102911.D13
Method     : C:\DX\METHOD\GLYCOLIC.MET
ACI Address: 1 System: 2 Inject#: 13              Detector: CDM-1
Analyst    : Ed Colvin                           Column: AG4A-SC,AS4A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1          606    4500    5Hz    0.00  15.00          0
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.98		0.000	67	488	1	
2	1.84		0.000	33	1268	1	
3	2.43	fluoride	16.304	25	122	2	0.00
4	2.71	acetate	2601.991	2141	20609	2	0.00
5	3.55	formate	979.093	1888	15873	1	0.00
6	4.47		0.000	369	3783	1	
7	5.70		0.000	32921	202446	2	
8	6.01		0.000	89616	1434364	2	
9	7.57		0.000	88188	3219882	2	

Totals 3597.388 215250 4898834

File: 98102911.D13 Sample: S98T002821 DUP

LABCORE Completed Worklist Report for Worklist# 26847

Analyst: adp

Instrument: IC40S1

Book#: 28N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0 @IC-QC F QC			1	<1.20e-2		ug/mL
1 CCB	0 @IC-QC CL QC			1	<1.70e-2		ug/mL
1 CCB	0 @IC-QC NO2 QC			1	<1.08e-1		ug/mL
1 CCB	0 @IC-QC BR QC			1	<1.25e-1		ug/mL
1 CCB	0 @IC-QC NO3 QC			1	<1.39e-1		ug/mL
1 CCB	0 @IC-QC P04 QC			1	<1.20e-1		ug/mL
1 CCB	0 @IC-QC S04 QC			1	<1.38e-1		ug/mL
1 CCB	0 @IC-QC OXALATE2 QC			1	<1.05e-1		ug/mL
2 ICV	0 @IC-QC F QC			5.90e1	5.85e+01	99.153 % Recovery	
2 ICV	0 @IC-QC CL QC			8.00e1	8.48e+01	106.000 % Recovery	
2 ICV	0 @IC-QC NO2 QC			5.70e2	5.83e+02	102.281 % Recovery	
2 ICV	0 @IC-QC BR QC			5.86e2	5.72e+02	97.611 % Recovery	
2 ICV	0 @IC-QC NO3 QC			5.92e2	5.84e+02	98.649 % Recovery	
2 ICV	0 @IC-QC P04 QC			5.47e2	5.40e+02	98.720 % Recovery	
2 ICV	0 @IC-QC S04 QC			6.38e2	6.49e+02	101.724 % Recovery	
2 ICV	0 @IC-QC OXALATE2 QC			5.44e2	5.59e+02	102.757 % Recovery	
3 CCV	0 @IC-QC F QC			6.40e1	5.85e+01	91.406 % Recovery	
3 CCV	0 @IC-QC CL QC			9.00e1	9.20e+01	102.222 % Recovery	
3 CCV	0 @IC-QC NO2 QC			5.53e2	5.61e+02	101.447 % Recovery	
3 CCV	0 @IC-QC BR QC			6.30e2	6.28e+02	99.683 % Recovery	
3 CCV	0 @IC-QC NO3 QC			6.98e2	7.08e+02	101.433 % Recovery	
3 CCV	0 @IC-QC P04 QC			6.32e2	6.36e+02	100.633 % Recovery	
3 CCV	0 @IC-QC S04 QC			6.99e2	7.10e+02	101.574 % Recovery	
3 CCV	0 @IC-QC OXALATE2 QC			5.47e2	5.73e+02	104.753 % Recovery	
4 SAMPLE	S98T003006 0 @IC-01 F-02 LIQUID			N/A	2.976e+02	61.810 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 CL-02 LIQUID			N/A	5.115e+03	87.570 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 NO2-02 LIQUID			N/A	6.965e+04	556.300 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 BR-02 LIQUID			N/A	7.135e+02	643.900 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 NO3-02 LIQUID			N/A	2.219e+05	716.000 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 P04-02 LIQUID			N/A	< 6.181e+02	618.100 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 S04-02 LIQUID			N/A	2.004e+03	710.800 ug/mL	
4 SAMPLE	S98T003006 0 @IC-01 OXALATE2 LIQUID			N/A	< 5.409e+02	540.900 ug/mL	
5 DUP	S98T003006 0 @IC-01 F-02 LIQUID			2.98e+02	2.80e+02	6.228 RPD	
5 DUP	S98T003006 0 @IC-01 CL-02 LIQUID			5.12e+03	5.00e+03	2.372 RPD	
5 DUP	S98T003006 0 @IC-01 NO2-02 LIQUID			6.96e+04	6.71e+04	3.658 RPD	
5 DUP	S98T003006 0 @IC-01 BR-02 LIQUID			7.14e+02	7.04e+02	1.410 RPD	
5 DUP	S98T003006 0 @IC-01 NO3-02 LIQUID			2.22e+05	2.20e+05	0.905 RPD	
5 DUP	S98T003006 0 @IC-01 P04-02 LIQUID			<6.18e2	<6.18e2	RPD	
5 DUP	S98T003006 0 @IC-01 S04-02 LIQUID			2.00e+03	2.02e+03	0.995 RPD	
5 DUP	S98T003006 0 @IC-01 OXALATE2 LIQUID			<5.41e2	<5.41e2	RPD	
6 SAMPLE	S98T003007 0 @IC-01 F-02 LIQUID			N/A	2.976e+02	61.810 ug/mL	

Units shown for QC (BLK/BKG) may not reflect the actual units.

1200

LABCORE Completed Worklist Report for Worklist# 26847

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6	SAMPLE	S98T003007	0	@IC-01	CL-02	LIQUID	N/A	4.641e+03	87.570 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	NO2-02	LIQUID	N/A	6.516e+04	556.300 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	BR-02	LIQUID	N/A	7.159e+02	643.900 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	NO3-02	LIQUID	N/A	2.012e+05	716.000 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	PO4-02	LIQUID	N/A	1.160e+03	618.100 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	SO4-02	LIQUID	N/A	2.126e+03	710.800 ug/mL
6	SAMPLE	S98T003007	0	@IC-01	OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL
7	DUP	S98T003007	0	@IC-01	F-02	LIQUID	2.98e+02	2.91e+02	2.377 RPD
7	DUP	S98T003007	0	@IC-01	CL-02	LIQUID	4.64e+03	4.68e+03	0.858 RPD
7	DUP	S98T003007	0	@IC-01	NO2-02	LIQUID	6.52e+04	6.55e+04	0.459 RPD
7	DUP	S98T003007	0	@IC-01	BR-02	LIQUID	7.16e+02	6.64e+02	7.536 RPD
7	DUP	S98T003007	0	@IC-01	NO3-02	LIQUID	2.01e+05	2.01e+05	0.000 RPD
7	DUP	S98T003007	0	@IC-01	PO4-02	LIQUID	1.16e+03	<6.18e2	RPD
7	DUP	S98T003007	0	@IC-01	SO4-02	LIQUID	2.13e+03	2.17e+03	1.860 RPD
7	DUP	S98T003007	0	@IC-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD

Final page for worklist# 26847

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11/2/98

LABCORE Data Entry Template for Worklist# 26847

Analyst: ADP Instrument: IC 4051 Book# 29N21-A ICV
Method: LA-533-105 Rev/Mod F-0 28N21-A CCV
Worklist Comment: BY-105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	ICV				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	SAMPLE	S98T003006 0	@IC-01		LIQUID		98000510	BY-105 (2)
		Analytes Requested:	BR-02		CL-02		F-02	NO2-02
			OXALATE2		PO4-02		SO4-02	NO3-02
5	DUP	S98T003006 0	@IC-01		LIQUID			
6	SAMPLE	S98T003007 0	@IC-01		LIQUID		98000510	BY-105 (2)
		Analytes Requested:	BR-02		CL-02		F-02	NO2-02
			OXALATE2		PO4-02		SO4-02	NO3-02
7	DUP	S98T003007 0	@IC-01		LIQUID			

Final page for worklist # 26847

10-29-98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

uploaded 10-29-98

Validated 11/6/98 JMS

JH Howell

268470CT.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-A ICV Date: 10/19/1998 00:55:38
 Data File : C:\DX\DATA\98102811.D21
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 21 Detector: CDM-1
 Analyst : *Anthony Perreault* Column: AG4A/AS4A anion column

10-29-98

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	78	314	2	
2	1.00	fluoride	58.489	99.13	10056	2	1.01
3	1.48	chloride	84.843	106.05	10503	1	-0.67
4	1.75	nitrite	582.936	102.27	49095	1	-2.05
5	2.61	bromide	571.654	97.55	34624	1	2.62
6	2.94	nitrate	583.791	98.61	47474	1	0.68
7	4.13	phosphate	539.603	98.65	19250	1	-1.59
8	5.33	sulfate	648.954	101.72	63284	1	-2.85
9	6.93	oxalate	558.702	102.70	40835	1	-3.03
Totals			3628.973	35322	275435		

File: 98102811.D21 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1203 TO 1209.

```

=====
Sample Name: 28N21-A CCV                      Date: 10/29/1998 02:20:37
Data File  : C:\DX\DATA\98102811.D22
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 22           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	79	297	2	
2	1.00	fluoride	58.499	1996	10058	2	1.01
3	1.48	chloride	92.027	2525	11417	1	-0.67
4	1.75	nitrite	560.644	8897	47179	1	-2.05
5	2.61	bromide	627.524	6312	38113	1	2.89
6	2.95	nitrate	708.141	8155	57924	1	0.91
7	4.13	phosphate	636.315	1789	22817	1	-1.75
8	5.28	sulfate	710.490	4633	69448	1	-3.83
9	6.93	oxalate	573.400	2610	41923	1	-3.03
Totals			3967.040	36998	299176		

File: 98102811.D22 Sample: 28N21-A CCV

```

=====
Sample Name: BLANK                      Date: 10/19/1998 00:41:32
Data File  : C:\DX\DATA\98102811.D20
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 20      Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    3000 5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	2.99	nitrate	0.104	18	93	1	2.28
Totals			0.104	18	93		

Sample Name: S98T003006 SAM Date: 10/29/1998 02:39:43
 Data File : C:\DX\DATA\98102811.D23
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 23 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	403	2216	2	
2	1.02	fluoride	297.615	253	1000	2	3.03
3	1.10		0.000	408	3517	2	
4	1.46	chloride	5115.451	2417	12473	1	-2.01
5	1.75	nitrite	69654.130	20543	117262	1	-2.05
6	2.61	bromide	713.509	63	321	1	2.89
7	2.85	nitrate	221868.190	45281	398759	1	-2.51
8	3.59		0.000	33	245	1	
9	4.13	phosphate	555.562	21	179	1	-1.59
10	5.33	sulfate	2004.446	254	3330	1	-2.85
Totals			300208.904	69675	539302		

File: 98102811.D23 Sample: S98T003006 SAM

```

=====
Sample Name: S98T003006 DUP                      Date: 10/29/1998 02:50:34
Data File  : C:\DX\DATA\98102811.D24
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 24              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   3000 5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	398	2094	2	
2	1.01	fluoride	280.064	253	942	2	2.36
3	1.09		0.000	400	3433	2	
4	1.45	chloride	5002.957	2470	12191	1	-2.46
5	1.74	nitrite	67117.048	20459	112877	1	-2.79
6	2.60	bromide	703.509	61	309	1	2.36
7	2.83	nitrate	219962.153	45218	394834	1	-2.97
8	3.59		0.000	33	243	1	
9	4.15	phosphate	552.970	19	177	1	-1.11
10	4.53		0.000	29	708	1	
11	5.33	sulfate	2015.201	268	3351	1	-2.85
12	6.99	oxalate	256.555	20	276	1	-2.28
Totals			295890.456	69628	531434		

File: 98102811.D24 Sample: S98T003006 DUP

```

=====
Sample Name: S98T003007 SAM                      Date: 10/19/1998 03:03:16
Data File  : C:\DX\DATA\98102811.D25
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 25              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000  5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	380	2083	2	
2	1.02	fluoride	297.577	249	1000	2	3.03
3	1.10		0.000	396	3327	2	
4	1.46	chloride	4641.066	2279	11286	1	-2.01
5	1.75	nitrite	65159.287	19619	109499	1	-2.42
6	2.62	bromide	715.946	62	324	1	3.15
7	2.85	nitrate	201242.706	41485	356867	1	-2.28
8	3.60		0.000	39	343	1	
9	4.14	phosphate	1159.510	37	599	1	-1.43
10	5.33	sulfate	2125.696	264	3560	1	-2.85
11	9.65		0.000	14	210	1	
Totals			275341.789	64823	489097		

File: 98102811.D25 Sample: S98T003007 SAM

Sample Name: S98T003007 DUP Date: 10/29/1998 03:14:47
 Data File : C:\DX\DATA\98102811.D26
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 26 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.0%		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	370	1975	2	
2	1.03	fluoride	291.045	247	978	2	3.70
3	1.10		0.000	402	3482	2	
4	1.47	chloride	4676.373	2349	11374	1	-1.57
5	1.75	nitrite	65487.340	19964	110064	1	-2.05
6	2.62	bromide	663.675	55	261	1	3.15
7	2.86	nitrate	201045.536	41520	356472	1	-2.05
8	3.60		0.000	19	127	1	
9	5.33	sulfate	2166.074	280	3637	1	-2.85
Totals			274330.043	65205	488372		

File: 98102811.D26 Sample: S98T003007 DUP

LABCORE Completed Worklist Report for Worklist# 26848

Analyst: adp

Instrument: IC40S1

Book#: 29N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC S04	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	ICV	0	@IC-QC F	QC	5.90e1	5.96e+01	101.017 % Recovery	
2	ICV	0	@IC-QC CL	QC	8.00e1	8.51e+01	106.375 % Recovery	
2	ICV	0	@IC-QC NO2	QC	5.70e2	5.91e+02	103.684 % Recovery	
2	ICV	0	@IC-QC BR	QC	5.86e2	5.72e+02	97.611 % Recovery	
2	ICV	0	@IC-QC NO3	QC	5.92e2	5.80e+02	97.973 % Recovery	
2	ICV	0	@IC-QC P04	QC	5.47e2	5.44e+02	99.452 % Recovery	
2	ICV	0	@IC-QC S04	QC	6.38e2	6.48e+02	101.567 % Recovery	
2	ICV	0	@IC-QC OXALATE2	QC	5.44e2	5.59e+02	102.757 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.11e+01	95.469 % Recovery	
3	CCV	0	@IC-QC CL	QC	9.00e1	9.08e+01	100.889 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.53e2	5.56e+02	100.542 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.07e+02	96.349 % Recovery	
3	CCV	0	@IC-QC NO3	QC	6.98e2	6.89e+02	98.711 % Recovery	
3	CCV	0	@IC-QC P04	QC	6.32e2	6.31e+02	99.842 % Recovery	
3	CCV	0	@IC-QC S04	QC	6.99e2	7.04e+02	100.715 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.47e2	5.53e+02	101.097 % Recovery	
4	SAMPLE	S98T003008	@IC-01 F-02	LIQUID	N/A	6.181e+01	61.810 ug/mL	
4	SAMPLE	S98T003008	@IC-01 CL-02	LIQUID	N/A	5.098e+03	87.570 ug/mL	
4	SAMPLE	S98T003008	@IC-01 NO2-02	LIQUID	N/A	7.270e+04	556.300 ug/mL	
4	SAMPLE	S98T003008	@IC-01 BR-02	LIQUID	N/A	6.439e+02	643.900 ug/mL	
4	SAMPLE	S98T003008	@IC-01 NO3-02	LIQUID	N/A	1.880e+05	716.000 ug/mL	
4	SAMPLE	S98T003008	@IC-01 P04-02	LIQUID	N/A	8.924e+02	618.100 ug/mL	
4	SAMPLE	S98T003008	@IC-01 S04-02	LIQUID	N/A	1.963e+03	710.800 ug/mL	
4	SAMPLE	S98T003008	@IC-01 OXALATE2	LIQUID	N/A	5.409e+02	540.900 ug/mL	
5	DUP	S98T003008	@IC-01 F-02	LIQUID	<6.18e-1	<6.18e-1	RPD	
5	DUP	S98T003008	@IC-01 CL-02	LIQUID	5.10e+03	5.44e+03	6.452 RPD	
5	DUP	S98T003008	@IC-01 NO2-02	LIQUID	7.27e+04	7.80e+04	7.034 RPD	
5	DUP	S98T003008	@IC-01 BR-02	LIQUID	<6.44e-2	<6.44e-2	RPD	
5	DUP	S98T003008	@IC-01 NO3-02	LIQUID	1.88e+05	1.97e+05	4.675 RPD	
5	DUP	S98T003008	@IC-01 P04-02	LIQUID	8.92e+02	8.73e+02	2.153 RPD	
5	DUP	S98T003008	@IC-01 S04-02	LIQUID	1.96e+03	2.17e+03	10.169 RPD	
5	DUP	S98T003008	@IC-01 OXALATE2	LIQUID	<5.41e-2	<5.41e-2	RPD	
6	SPK	S98T003008	@IC-01 F-02	LIQUID	5.90e1	4.95e+01	83.898 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26848

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SPK	S98T003008	0	⊗IC-01	CL-02	LIQUID	8.00e1	7.51e+01	93.875 % Recovery
6 SPK	S98T003008	0	⊗IC-01	NO2-02	LIQUID	5.70e2	5.58e+02	97.895 % Recovery
6 SPK	S98T003008	0	⊗IC-01	BR-02	LIQUID	5.85e2	5.60e+02	95.563 % Recovery
6 SPK	S98T003008	0	⊗IC-01	NO3-02	LIQUID	5.92e2	5.10e+02	86.149 % Recovery
6 SPK	S98T003008	0	⊗IC-01	PO4-02	LIQUID	5.47e2	5.37e+02	98.172 % Recovery
6 SPK	S98T003008	0	⊗IC-01	SO4-02	LIQUID	6.38e2	6.36e+02	99.687 % Recovery
6 SPK	S98T003008	0	⊗IC-01	OXALATE2	LIQUID	5.44e2	5.53e+02	101.654 % Recovery
7 SAMPLE	S98T003009	0	⊗IC-01	F-02	LIQUID	N/A	< 6.181e+01	61.810 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	CL-02	LIQUID	N/A	5.074e+03	87.570 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	NO2-02	LIQUID	N/A	7.166e+04	556.300 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	BR-02	LIQUID	N/A	< 6.435e+02	643.900 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	NO3-02	LIQUID	N/A	1.941e+05	716.000 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	PO4-02	LIQUID	N/A	7.276e+02	618.100 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	SO4-02	LIQUID	N/A	2.076e+03	710.800 ug/mL
7 SAMPLE	S98T003009	0	⊗IC-01	OXALATE2	LIQUID	N/A	< 5.409e+02	540.900 ug/mL
8 DUP	S98T003009	0	⊗IC-01	F-02	LIQUID	<6.18e1	<6.18e1	RPD
8 DUP	S98T003009	0	⊗IC-01	CL-02	LIQUID	5.07e+03	5.14e+03	1.371 RPD
8 DUP	S98T003009	0	⊗IC-01	NO2-02	LIQUID	7.17e+04	7.28e+04	1.522 RPD
8 DUP	S98T003009	0	⊗IC-01	BR-02	LIQUID	<6.44e2	<6.44e2	RPD
8 DUP	S98T003009	0	⊗IC-01	NO3-02	LIQUID	1.94e+05	1.99e+05	2.545 RPD
8 DUP	S98T003009	0	⊗IC-01	PO4-02	LIQUID	7.28e+02	9.89e+02	30.402 RPD
8 DUP	S98T003009	0	⊗IC-01	SO4-02	LIQUID	2.08e+03	2.13e+03	2.375 RPD
8 DUP	S98T003009	0	⊗IC-01	OXALATE2	LIQUID	<5.41e2	<5.41e2	RPD

Final page for worklist# 26848

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jane M. Lacy
Reviewer Signature _____ Date 11/2/98

LABCORP Data Entry Template for Worklist# 26848

Analyst: ADP Instrument: IC 4051 Book# 29N21-A ICV
Method: LA-533-105 Rev/Mod F-0 28N21-A CCV

Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R #	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 ICV			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 SAMPLE	S98T003008 0		@IC-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested:		BR-02	CL-02	F-02	NO2-02
			OXALATE2	PO4-02	SO4-02	NO3-02
5 DUP	S98T003008 0		@IC-01	LIQUID		
6 SPK	S98T003008 0		@IC-01	LIQUID		
7 SAMPLE	S98T003009 0		@IC-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested:		BR-02	CL-02	F-02	NO2-02
			OXALATE2	PO4-02	SO4-02	NO3-02
8 DUP	S98T003009 0		@IC-01	LIQUID		

Final page for worklist # 26848

Anthony Pimenton 10-27-98
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

uploaded 10-29-98

Validated 11/2/98 JMD

JH Worell

2684800T.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-A ICV Date: 10/27/1998 19:43:49
 Data File : C:\DX\DATA\98102711.D13
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: A System: 1 Inject#: 13 Detector: CDM-1
 Analyst : *Anthony Kuntz* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	71	274	2	
2	1.01	fluoride	59.631	101.07	10255	2	1.68
3	1.47	chloride	85.125	106.41	10538	1	-1.12
4	1.75	nitrite	591.227	103.72	49808	1	-2.05
5	2.58	bromide	572.022	97.61	34647	1	1.57
6	2.91	nitrate	580.048	97.98	47161	1	-0.23
7	4.21	phosphate	543.935	94.34	19410	1	0.32
8	5.44	sulfate	648.146	101.59	63203	1	-0.91
9	7.09	oxalate	558.713	102.70	40836	1	-0.79
Totals			3638.846	35268	276131		

File: 98102711.D13 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1213 TO 1220.

1213

Sample Name: 28N21-A CCV Date: 10/27/1998 20:00:38
 Data File : C:\DX\DATA\98102711.D14
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 14 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	76	293	2	
2	1.00	fluoride	61.095	2139	10510	2	1.01
3	1.47	chloride	90.822	2547	11263	1	-1.57
4	1.74	nitrite	556.053	9200	46785	1	-2.79
5	2.57	bromide	607.285	6485	36848	1	1.05
6	2.89	nitrate	689.404	8371	56346	1	-0.91
7	4.19	phosphate	631.294	1827	22631	1	-0.32
8	5.44	sulfate	703.523	5650	68749	1	-0.91
9	7.04	oxalate	552.522	2221	40377	1	-1.54
Totals			3891.998	38515	293803		

File: 98102711.D14 Sample: 28N21-A CCV

1214

Sample Name: BLANK Date: 10/27/1998 19:27:13
Data File : C:\DX\DATA\98102711.D12
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 12 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.39	chloride	0.012	9	37	1	-6.49
Totals			0.012	9	37		

1215

Sample Name: S98T003008 SAM Date: 10/27/1998 21:46:44
 Data File : C:\DX\DATA\98102711.D17
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 17 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	432	2454	2	
2	1.10		0.000	415	4188	2	
3	1.46	chloride	5097.939	2509	12429	1	-2.01
4	1.75	nitrite	72696.986	22456	122530	1	-2.42
5	2.59	bromide	544.501	25	120	1	1.84
6	2.83	nitrate	187986.105	39793	330583	1	-3.20
7	3.73		0.000	37	309	1	
8	4.21	phosphate	892.427	40	413	1	0.32
9	5.44	sulfate	1963.429	227	3253	1	-0.91
10	7.20	oxalate	315.353	18	359	1	0.70
Totals			269496.740	65951	476638		

File: 98102711.D17 Sample: S98T003008 SAM

Sample Name: S98T003008 DUP Date: 10/27/1998 22:07:47
 Data File : C:\DX\DATA\98102711.D18
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	461	2660	2	
2	1.10		0.000	456	4651	2	
3	1.46	chloride	5440.965	2636	13290	1	-2.01
4	1.75	nitrite	78037.626	23515	131803	1	-2.42
5	2.59	bromide	522.900	20	94	1	1.84
6	2.83	nitrate	197320.148	41756	349039	1	-3.20
7	3.72		0.000	37	308	1	
8	4.20	phosphate	873.428	38	400	1	0.00
9	5.44	sulfate	2169.606	257	3644	1	-0.91
Totals			284364.673	69176	505888		

File: 98102711.D18 Sample: S98T003008 DUP

1217

Data Reprocessed On 10/29/1998 08:31:04

```

=====
Sample Name: S98T003008 SPK                      Date: 10/27/1998 22:20:09
Data File  : F:\DATA\98102711.D19
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 19              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151    3000    5Hz    0.00 10.00          30
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	439	1957	2	
2	0.99	fluoride	2502.041	1399	8418	2	0.34
3	1.44	chloride	8891.093	4538	22082	1	-3.36
4	1.73	nitrite	100891.264	31371	171861	1	-3.54
5	2.54	bromide	28260.820	6039	33532	1	0.00
6	2.79	nitrate	213727.736	45496	382075	1	-0.12
7	3.67		0.000	32	233	1	
8	4.14	phosphate	28021.688	1574	19612	1	-1.43
9	5.39	sulfate	34064.830	5416	65183	1	-1.88
10	6.99	oxalate	27933.496	2296	40022	1	-2.28
Totals			444292.970	98600	744975		

File: 98102711.D19 Sample: S98T003008 SPK

=====

Sample Name: S98T003009 SAM	Date: 10/27/1998 22:31:53
Data File : C:\DX\DATA\98102711.D20	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 20	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	417	2399	2	
2	1.10		0.000	419	4317	2	
3	1.46	chloride	5073.641	2460	12368	1	-2.01
4	1.75	nitrite	71658.171	21758	120730	1	-2.42
5	2.59	bromide	587.701	33	171	1	1.84
6	2.83	nitrate	194137.967	41503	342720	1	-3.20
7	3.71		0.000	47	467	1	
8	4.19	phosphate	727.612	30	299	1	-0.32
9	5.44	sulfate	2076.249	258	3467	1	-0.91
Totals			274261.341	66924	486938		

File: 98102711.D20 Sample: S98T003009 SAM

Sample Name: S98T003009 DUP Date: 10/27/1998 22:45:43
 Data File : C:\DX\DATA\98102711.D21
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 21 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	439	2512	2	
2	1.09		0.000	417	4118	2	
3	1.45	chloride	5136.590	2467	12526	1	-2.46
4	1.73	nitrite	72797.782	21973	122705	1	-3.17
5	2.57	bromide	595.937	37	181	1	1.31
6	2.81	nitrate	198645.179	41827	351679	1	-3.65
7	3.70		0.000	30	219	1	
8	4.19	phosphate	989.428	43	480	1	-0.32
9	5.44	sulfate	2130.967	280	3570	1	-0.91
10	7.09	oxalate	320.909	23	367	1	-0.79
Totals			280616.793	67537	498357		

File: 98102711.D21 Sample: S98T003009 DUP

LABCORE Completed Worklist Report for Worklist# 26849

Analyst: adp

Instrument: IC45S2

Book#: 34N21A

Method: LA-533-105

Rev/Mod F-D

Worklist Comment: BY-015 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	5.89e1	5.98e+01	101.528 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.24e2	1.09e+02	87.903 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.08e+02	90.756 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLTI	QC	1.07e2	9.78e+01	91.402 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.21e1	5.91e+01	95.169 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.13e+02	91.870 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.21e+02	93.077 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLTI	QC	1.02e2	9.51e+01	93.235 % Recovery	
5 DUP	S98T003006 0	@IC4G-01 F*4-01	LIQUID	5.57e+02	2.53e+02	75.062 RPD	
5 DUP	S98T003006 0	@IC4G-01 ACETATE2	LIQUID	8.91e+02	1.75e+03	65.051 RPD	
5 DUP	S98T003006 0	@IC4G-01 FORMATE2	LIQUID	7.05e+02	1.22e+03	53.506 RPD	
5 DUP	S98T003006 0	@IC4G-01 GLYCOLTI	LIQUID	1.27e+03	<1.19e2	RPD	
6 SAMPLE	S98T003007 0	@IC4G-01 F*4-01	LIQUID	N/A	2.251e+02	65.750 ug/mL	
6 SAMPLE	S98T003007 0	@IC4G-01 ACETATE2	LIQUID	N/A	1.642e+03	123.000 ug/mL	
6 SAMPLE	S98T003007 0	@IC4G-01 FORMATE2	LIQUID	N/A	1.134e+03	131.500 ug/mL	
6 SAMPLE	S98T003007 0	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 1.188e+02	118.800 ug/mL	
7 DUP	S98T003007 0	@IC4G-01 F*4-01	LIQUID	2.25e+02	2.28e+02	1.325 RPD	
7 DUP	S98T003007 0	@IC4G-01 ACETATE2	LIQUID	1.64e+03	1.66e+03	1.212 RPD	
7 DUP	S98T003007 0	@IC4G-01 FORMATE2	LIQUID	1.13e+03	1.14e+03	0.881 RPD	
7 DUP	S98T003007 0	@IC4G-01 GLYCOLTI	LIQUID	<1.19e2	<1.19e2	RPD	

Final page for worklist# 26849

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jean M. Luy 1/11/99
Reviewer Signature Date

Sample # S98T003006 was reanalyzed on WL # 27933.
J. Luy 3/23/96

11/05/98 14:04
A-0004-1

Page: 1.

LABCORE Data Entry Template for Worklist# 26849

Analyst: ADP Instrument: IC _____ Book# 35N21-C LCS
34N21-A CCVMethod: LA-533-1105 Rev/Mod F-0

Worklist Comment: BY-015 FOR @IC4G-01 RTS

105 TMS 11/20/98

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T003006 0		@IC4G-01 LIQUID		98000510	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
5 DUP	S98T003006 0		@IC4G-01 LIQUID			
6 SAMPLE	S98T003007 0		@IC4G-01 LIQUID		98000510	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
7 DUP	S98T003007 0		@IC4G-01 LIQUID			

Final page for worklist # 26849

Anthony P. Smith 12-31-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

*Note Blank Dilution is 1

uploaded 1-4-99
Sh. Howell 26849 DEC. CSV

S98T003006 was rejected
for high RPD.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Rest Validated 1/11/98 gmd

Data Reprocessed On 12/31/1998 15:16:50

Sample Name: 35N21-C LCS Date: 12/31/1998 15:13:40
 Data File : C:\DX\DATA\98123101.D16
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 3 System: 2 Inject#: 16 Detector: CDM-2
 Analyst: *Anthony Purnell* Column: AG14A-SC, AS14A-SC, SRS

12-31-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External	1	101	5700	5Hz	0.00	19.00	16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	5544	152888	1	
5	2.87		0.000	2598	26492	2	
6	3.04		0.000	2500	37493	2	
7	3.33		0.000	2662	31064	2	
26	7.17	fluoride	59.826101	571803656	25659223	2	0.00
27	7.89	glycolate	97.8429144	502450	8709627	2	0.00
28	8.37	acetate	108.95587	572513	10918751	2	0.00
29	9.07	formate	108.21290	9369671	18154780	3	0.00
31	11.33		0.000	8631952	383238456	2	
32	12.05		0.000	7463263	618686083	2	
33	17.33		0.000	940371	49531020	1	
Totals			374.834	20897181	1115145878		

File: 98123101.D16 Sample: 35N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1223 TO 1226.

Data Reprocessed On 12/31/1998 15:36:05

```

=====
Sample Name: 34N21-A CCV                               Date: 12/31/1998 15:33:45
Data File  : C:\DX\DATA\98123101.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17                    Detector: CDM-2
Analyst    :                                           Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz   0.00 19.00   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	3524	53271	2	
5	2.24		0.000	2405	23196	1	
6	2.76		0.000	2188	32598	2	
7	2.95		0.000	2307	17503	2	
12	3.80		0.000	3623	55731	2	
13	4.03		0.000	4541	28783	2	
14	4.12		0.000	5087	68497	2	
15	4.37		0.000	5424	31130	2	
16	4.65		0.000	8460	121529	2	
17	4.84		0.000	9251	185857	2	
22	5.72		0.000	1774	20144	2	
23	5.95		0.000	1497	23965	2	
26	7.15	fluoride	59.070	1780694	25325569	2	0.00
27	7.87	glycolate	95.052	489752	8460781	2	0.00
28	8.35	acetate	113.198	594590	11333739	2	0.00
29	9.04	formate	120.655	1088901	20281429	2	0.00
30	11.33		0.000	8694324	398835388	2	
31	12.05		0.000	7396881	670118864	2	
Totals			387.975	20095222	1135017975		

Data Reprocessed On 01/04/1999 13:38:31

```

=====
Sample Name: BLANK                      Date: 12/31/1998 14:53:46
Data File  : E:\DATA\98123101.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 15    Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          1  5700  5Hz  0.00 19.00  16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.47		0.000	8508	413608	1	
3	2.65		0.000	1740	16304	3	
21	7.31	fluoride	0.009	629	37160	1	0.00
22	11.33		0.000	8910104	504564867	2	
23	12.05		0.000	8151084	647348944	3	
24	17.73		0.000	5973	148038	4	
25	18.43		0.000	1848	27478	4	
Totals			0.009	17079886	1152556399		

File: 98123101.D15 Sample: BLANK

Data Reprocessed On 01/04/1999 13:40:59

```

=====
Sample Name: S98T003007 SAM                      Date: 12/31/1998 16:34:48
Data File  : E:\DATA\98123101.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          2121   5700  5Hz   6.58 10.05      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	43160	402784	1	
5	2.10		0.000	2189	25080	1	
6	2.89		0.000	7228	154436	2	
7	3.01		0.000	6599	123487	2	
8	3.40		0.000	2715	18868	3	
23	5.32		0.000	4315	41701	2	
24	5.51		0.000	5944	275891	2	
25	7.09	fluoride	225.082	257885	4251299	2	0.00
26	7.84	glycolate	76.390	34449	480069	2	0.00
27	8.29	acetate	1642.493	399400	7908819	2	0.00
28	8.99	formate	1133.698	482204	8968901	2	0.00
29	9.76		0.000	88563	1746077	2	
30	10.88		0.000	12063455	95819256	2	
31	11.17		0.000	73142762	733181240	2	
32	11.84		0.000	156040277	2358673571	2	
Totals			3077.662	242581147	3212071479		

File: 98123101.D20 Sample: S98T003007 SAM

```

=====
Sample Name: S98T003007 DUP                      Date: 12/31/1998 16:54:53
Data File  : E:\DATA\98123101.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 21              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 2121 5700 5Hz 6.37 10.86 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	6155	164124	1	
3	1.73		0.000	45147	429024	1	
5	3.00		0.000	14016	401910	2	
6	3.25		0.000	14692	114128	2	
7	3.36		0.000	14851	73627	2	
8	3.45		0.000	15018	143753	2	
9	3.68		0.000	14549	249454	2	
10	3.93		0.000	14164	80532	2	
11	4.03		0.000	14031	123931	2	
12	4.23		0.000	14677	212994	2	
13	4.47		0.000	14541	125975	2	
14	4.63		0.000	14269	93409	2	
15	4.79		0.000	14998	143972	2	
16	4.92		0.000	15602	135264	2	
17	5.11		0.000	15456	310028	2	
18	5.64		0.000	15392	612374	2	
19	7.12	fluoride	227.825	259477	4307521	2	0.00
20	7.87	glycolate	70.003	33951	454052	2	0.00
21	8.32	acetate	1661.427	404559	7997191	2	0.00
22	9.01	formate	1143.421	485913	9046779	2	0.00
23	9.79		0.000	88193	1716787	2	
24	10.88		0.000	12364390	94984438	2	
25	11.15		0.000	64287901	734134841	2	
26	11.84		0.000	156643029	2376182525	2	

Totals 3102.676 234824973 3232238632

LABCORE Completed Worklist Report for Worklist# 26850

Analyst: vlm

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-015 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 ICV	0	@IC4G-QC F*4	QC	5.89e1	6.14e+01	104.244 % Recovery	
2 ICV	0	@IC4G-QC ACETATE2	QC	1.24e2	1.15e+02	92.742 % Recovery	
2 ICV	0	@IC4G-QC FORMATE2	QC	1.19e2	1.11e+02	93.277 % Recovery	
2 ICV	0	@IC4G-QC GLYCOLTI	QC	1.07e2	1.02e+02	95.327 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.21e1	6.00e+01	96.618 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.17e+02	95.122 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.24e+02	95.385 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLTI	QC	1.02e2	9.98e+01	97.843 % Recovery	
4 SAMPLE	S98T003008	@IC4G-01 F*4-01	LIQUID	N/A	<1.597e+02	159.700	ug/mL
4 SAMPLE	S98T003008	@IC4G-01 ACETATE2	LIQUID	N/A	1.496e+03	298.800	ug/mL
4 SAMPLE	S98T003008	@IC4G-01 FORMATE2	LIQUID	N/A	1.019e+03	319.400	ug/mL
4 SAMPLE	S98T003008	@IC4G-01 GLYCOLTI	LIQUID	N/A	<2.885e+02	288.500	ug/mL
5 DUP	S98T003008	@IC4G-01 F*4-01	LIQUID	<1.60e2	<1.60e2		RPD
5 DUP	S98T003008	@IC4G-01 ACETATE2	LIQUID	1.50e+03	1.89e+03	23.009	RPD
5 DUP	S98T003008	@IC4G-01 FORMATE2	LIQUID	1.02e+03	1.16e+03	12.844	RPD
5 DUP	S98T003008	@IC4G-01 GLYCOLTI	LIQUID	<2.88e2	<2.88e2		RPD
6 SPK	S98T003008	@IC4G-01 F*4-01	LIQUID	5.89e1	6.36e+01	107.980 % Recovery	
6 SPK	S98T003008	@IC4G-01 ACETATE2	LIQUID	1.24e2	1.25e+02	100.806 % Recovery	
6 SPK	S98T003008	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.21e+02	101.691 % Recovery	
6 SPK	S98T003008	@IC4G-01 GLYCOLTI	LIQUID	1.07e2	1.10e+02	102.804 % Recovery	
7 SAMPLE	S98T003009	@IC4G-01 F*4-01	LIQUID	N/A	1.690e+02	159.700	ug/mL
7 SAMPLE	S98T003009	@IC4G-01 ACETATE2	LIQUID	N/A	1.764e+03	298.800	ug/mL
7 SAMPLE	S98T003009	@IC4G-01 FORMATE2	LIQUID	N/A	1.295e+03	319.400	ug/mL
7 SAMPLE	S98T003009	@IC4G-01 GLYCOLTI	LIQUID	N/A	<2.885e+02	288.500	ug/mL
8 DUP	S98T003009	@IC4G-01 F*4-01	LIQUID	1.69e+02	1.80e+02	6.394	RPD
8 DUP	S98T003009	@IC4G-01 ACETATE2	LIQUID	1.76e+03	1.59e+03	10.149	RPD
8 DUP	S98T003009	@IC4G-01 FORMATE2	LIQUID	1.30e+03	1.26e+03	4.511	RPD
8 DUP	S98T003009	@IC4G-01 GLYCOLTI	LIQUID	<2.88e2	<2.88e2		RPD

Final page for worklist# 26850

Analyst Signature Date

Analyst Signature Date

James M. Lyle 1/15/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26850

Analyst: VLM Instrument: IC ☒

Book# CCV 34 N21-A

Method: LA-533 ^{gmp} 105 Rev/Mod E-0

Worklist Comment: BY-015 FOR @IC4G-01 RTS

(has been last
resend)

105 TMS 10/27/98

S Type	Sample#	R A	Test	Matrix	Group#	Project

1	CCB		@IC4G-QC	QC		

2	ICV		@IC4G-QC	QC		

3	CCV		@IC4G-QC	QC		

4	SAMPLE	S98T003008	0	@IC4G-01 LIQUID	98000510	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						

5	DUP	S98T003008	0	@IC4G-01 LIQUID		

6	SPK	S98T003008	0	@IC4G-01 LIQUID		

7	SAMPLE	S98T003009	0	@IC4G-01 LIQUID	98000510	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						

8	DUP	S98T003009	0	@IC4G-01 LIQUID		

Final page for worklist # 26850

Valent Masari 01-13-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

up loaded 1-14-99

Validated 1/14/99 M. J. J.

Dr. Worell
26850.JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 01/13/1999 08:54:56
 Data File : C:\DX\DATA\99011301.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst : *Valerie L. Masia* Column: AG14A-SC, AS14A-SC, SRS
01-13-99

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	29195	790628	1	
2	1.63		0.000	20089	235824	1	
4	5.76		0.000	100432	24250106	2	
5	5.88		0.000	42173	474916	2	
6	6.72		0.000	18811	4567230	1	
7	11.33		0.000	7600511	321469432	2	
8	11.73		0.000	6573830	70038877	2	
9	12.03		0.000	11239397	654928848	3	
10	17.81		0.000	4614	97538	4	
Totals			0.000	25629050	1076853399		

File: 99011301.D01 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES *1231* TO *1242*.

Sample Name: LCS 35N21-C Date: 01/13/1999 09:17:24
 Data File : C:\DX\DATA\99011301.D02
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2
 Analyst : *Valerie Macal* Column: AG14A-SC, AS14A-SC, SRS

01-13-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	18225	927620	1	
5	2.88		0.000	3659	37427	2	
6	2.97		0.000	4757	58084	2	
8	3.43		0.000	1887	28343	2	
10	3.92		0.000	3272	39812	2	
11	4.15		0.000	2772	41644	2	
22	7.31	fluoride	104.23	61.394	1753829	2	0.00
23	8.03	glycolate	95.77	102.472	503085	2	0.00
24	8.53	acetate	92.69	114.934	587951	2	0.00
25	9.25	formate	92.95	110.615	999007	2	0.00
26	11.36		0.000	7869306	365177622	2	
27	12.03		0.000	8925546	712774174	3	
28	17.89		0.000	1698	20486	4	
Totals			389.415	20674994	1144649156		

File: 99011301.D02 Sample: LCS 35N21-C

Sample Name: CCV 34N21-A Date: 01/13/1999 09:41:54
 Data File : C:\DX\DATA\99011301.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3
 Analyst : *Valerie Mason* Column: AG14A-SC, AS14A-SC, SRS
 Detector: CDM-2

01-13-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	10807	354836	2	
6	2.77		0.000	3109	24937	2	
7	2.96		0.000	3056	39302	2	
8	3.20		0.000	2565	28674	2	
10	3.63		0.000	1640	20906	2	
13	4.23		0.000	3262	46124	2	
14	4.53		0.000	3734	59571	2	
15	5.00		0.000	2931	17770	2	
16	5.12		0.000	2715	17928	2	
19	5.49		0.000	1449	17310	2	
21	6.07		0.000	1720	20462	2	
24	7.31	fluoride	96.69 60.045	1741632	25755924	2	0.00
25	8.03	glycolate	97.85 99.805	488085	8884880	2	0.00
26	8.53	acetate	95.13 117.014	612127	11706864	2	0.00
27	9.23	formate	95.07 123.594	1109242	20784827	2	0.00
28	11.33		0.000	8137692	374013020	2	
29	12.03		0.000	8947194	710511465	3	
30	17.52		0.000	5562	221129	4	
Totals			400.457	21078522	1152525927		

File: 99011301.D03 Sample: CCV 34N21-A

1233

Data Reprocessed On 01/14/1999 10:29:37

```

=====
Sample Name: S98T003008                      Date: 01/13/1999 10:53:07
Data File  : E:\DATA\99011311.D02
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 2           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      5151   5700   5Hz   6.79  10.55      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	11588	175620	2	
2	0.63		0.000	4001	54172	2	
5	2.23		0.000	7200	92360	1	
6	2.67		0.000	4968	31864	2	
7	2.75		0.000	7211	41147	2	
8	2.87		0.000	6376	31401	2	
9	3.00		0.000	7182	38030	2	
10	3.16		0.000	5269	30654	2	
11	3.23		0.000	8872	46632	3	
14	3.59		0.000	2737	17406	2	
16	3.92		0.000	3984	27666	2	
17	4.08		0.000	4495	36349	2	
18	4.20		0.000	5328	97108	2	
19	4.59		0.000	7679	119713	2	
20	4.89		0.000	9874	67884	2	
21	5.04		0.000	9559	65620	2	
22	5.27		0.000	10199	119904	2	
23	5.53		0.000	9917	101361	2	
24	5.69		0.000	7828	29966	2	
25	5.80		0.000	9635	226018	2	
26	6.61		0.000	3134	20086	2	
27	6.77		0.000	3352	28218	2	
29	7.31	fluoride	101.972	36339	507600	1	0.00
30	8.08	glycolate	19.376	16969	201578	2	0.00
31	8.53	acetate	1495.658	149662	3109508	2	0.00
32	9.23	formate	1018.857	162262	3274339	2	0.00
33	10.03		0.000	40954	795508	2	
34	10.88		0.000	5523693	59194991	2	
35	11.17		0.000	36732947	467415690	2	
36	11.89		0.000	61713570	1073685608	3	
37	18.16		0.000	2328	24675	4	
38	18.43		0.000	2220	30384	4	

```

-----
Totals                2635.863  104531331  1609739058

```

Data Reprocessed On 01/14/1999 10:31:47

```

=====
Sample Name: S98T003008 DUP                      Date: 01/13/1999 11:14:46
Data File  : E:\DATA\99011311.D03
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3                Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          5151   5700  5Hz   6.37 10.39    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	4400	85920	1	
5	2.25		0.000	6600	84480	1	
6	2.69		0.000	3000	24722	2	
11	3.21		0.000	2316	16402	2	
15	3.87		0.000	2837	31624	2	
16	4.27		0.000	2479	26080	2	
19	4.69		0.000	2884	19973	2	
20	4.87		0.000	2242	17373	2	
21	5.01		0.000	2837	17484	2	
22	5.23		0.000	2540	42077	2	
25	5.84		0.000	1466	30080	1	
26	6.77		0.000	6285	111443	2	
27	7.28	fluoride	141.837	42700	842550	2	0.00
28	8.48	acetate	1891.786	154974	3873059	2	0.00
29	9.20	formate	1155.926	173138	3722556	2	0.00
30	10.03		0.000	47326	1115203	2	
31	10.88		0.000	5691520	59538015	2	
32	11.17		0.000	37222587	468900137	2	
33	11.89		0.000	62302751	1078836334	2	
Totals			3189.548	105674882	1617335513		

Data Reprocessed On 01/14/1999 10:35:31

Sample Name: S98T003008 SPK	Date: 01/13/1999 11:36:25
Data File : E:\DATA\99011301.D04	
Method : C:\DX\METHOD\AAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 4	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	5700	5Hz	5.93	11.06	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	3732	72240	1	
3	1.27		0.000	3000	56760	1	
4	2.27		0.000	6600	71120	1	
5	2.87		0.000	5709	95220	2	
6	3.05		0.000	4428	51611	2	
7	3.16		0.000	4525	28034	2	
8	3.35		0.000	2644	26672	2	
15	4.81		0.000	2696	43439	2	
16	5.05		0.000	3254	23512	2	
17	5.36		0.000	13911	541753	2	
18	7.31	fluoride	3212.256	1863476	27056038	2	0.00
19	8.03	glycolate	5560.246	543788	9710258	2	0.00
20	8.51	acetate	7831.565	765996	15275453	2	0.00
21	9.23	formate	7104.790	1244308	23483246	3	0.00
22	10.00		0.000	51153	1011884	4	
23	10.88		0.000	5657025	59545087	2	
24	11.17		0.000	37004152	468425903	2	
25	11.89		0.000	61964465	1078334677	2	

Totals	23708.857	109144862	1683852906

```

=====
Sample Name: S98T003009                      Date: 01/13/1999 11:58:02
Data File  : E:\DATA\99011301.D05
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5           Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700  5Hz   6.24 10.52      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	11973	487457	1	
2	2.27		0.000	7000	86105	1	
3	2.96		0.000	3927	82847	3	
9	3.84		0.000	2565	23055	1	
12	4.35		0.000	886	16369	1	
13	4.92		0.000	4433	53226	2	
14	5.16		0.000	5513	38136	2	
15	5.35		0.000	14130	476704	2	
16	6.01		0.000	8907	132136	2	
17	6.64		0.000	3054	17777	2	
18	6.85		0.000	4730	60786	2	
19	7.33	fluoride	168.957	61401	1070496	2	0.00
20	8.24	glycolate	246.174	27226	581839	2	0.00
21	8.53	acetate	1763.772	171830	3626352	2	0.00
22	9.25	formate	1294.766	198088	4176881	2	0.00
23	10.05		0.000	54374	821147	2	
24	10.88		0.000	5984669	62949927	2	
25	11.17		0.000	40345774	502406194	2	
26	11.89		0.000	69486031	1183331571	2	
Totals			3473.669	116396510	1760439004		

```

=====
Sample Name: S98T003009 DUP                      Date: 01/13/1999 12:19:40
Data File  : E:\DATA\99011301.D06
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6                Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700  5Hz   6.29 10.28      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	8400	272160	1	
3	2.27		0.000	6800	95080	1	
4	2.92		0.000	6508	137055	2	
5	3.08		0.000	5529	69213	2	
6	3.27		0.000	5254	18627	2	
7	3.35		0.000	4965	25393	2	
8	3.45		0.000	3379	21764	2	
15	5.28		0.000	3808	31232	2	
16	5.41		0.000	3666	101441	2	
17	5.91		0.000	5362	118136	2	
18	7.31	fluoride	180.175	62038	1164806	2	0.00
19	8.15	glycolate	169.070	24508	452480	2	0.00
20	8.51	acetate	1589.241	172562	3289928	2	0.00
21	9.23	formate	1363.581	200876	4402181	2	0.00
22	10.00		0.000	53214	1102036	2	
23	10.88		0.000	6275599	63239259	2	
24	11.17		0.000	40986727	495790243	2	
25	11.87		0.000	64127029	1155217456	3	
26	18.24		0.000	3222	73790	4	
Totals			3302.066	111959446	1725622280		

LABCORE Completed Worklist Report for Worklist# 26988

Analyst: adp

Instrument: IC40S1

Book#: 28N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC PO4	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC SO4	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	ICV	0	@IC-QC F	QC	5.90e1	5.94e+01	100.678 % Recovery	
2	ICV	0	@IC-QC CL	QC	8.00e1	8.82e+01	110.250 % Recovery	
2	ICV	0	@IC-QC NO2	QC	5.70e2	5.66e+02	99.298 % Recovery	
2	ICV	0	@IC-QC BR	QC	5.86e2	5.71e+02	97.440 % Recovery	
2	ICV	0	@IC-QC NO3	QC	5.92e2	5.93e+02	100.169 % Recovery	
2	ICV	0	@IC-QC PO4	QC	5.47e2	5.49e+02	100.366 % Recovery	
2	ICV	0	@IC-QC SO4	QC	6.38e2	6.54e+02	102.508 % Recovery	
2	ICV	0	@IC-QC OXALATE2	QC	5.44e2	5.53e+02	101.654 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.08e+01	95.000 % Recovery	
3	CCV	0	@IC-QC CL	QC	9.00e1	9.35e+01	103.889 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.53e2	5.50e+02	99.458 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.10e+02	96.825 % Recovery	
3	CCV	0	@IC-QC NO3	QC	6.98e2	7.16e+02	102.579 % Recovery	
3	CCV	0	@IC-QC PO4	QC	6.32e2	6.48e+02	102.532 % Recovery	
3	CCV	0	@IC-QC SO4	QC	6.99e2	6.93e+02	99.142 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.47e2	5.60e+02	102.377 % Recovery	
4	SAMPLE	S98T002964	0	@IC-01 F-02	LIQUID	N/A	< 4.969e+01	49.690 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 CL-02	LIQUID	N/A	5.224e+03	70.400 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 NO2-02	LIQUID	N/A	7.663e+04	447.200 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 BR-02	LIQUID	N/A	< 5.176e+02	517.600 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 NO3-02	LIQUID	N/A	1.740e+05	575.600 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 PO4-02	LIQUID	N/A	9.228e+02	496.900 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 SO4-02	LIQUID	N/A	1.691e+03	571.500 ug/mL
4	SAMPLE	S98T002964	0	@IC-01 OXALATE2	LIQUID	N/A	< 4.348e+02	434.800 ug/mL
5	DUP	S98T002964	0	@IC-01 F-02	LIQUID	<4.97e1	<4.97e1	RPD
5	DUP	S98T002964	0	@IC-01 CL-02	LIQUID	5.22e+03	5.33e+03	2.085 RPD
5	DUP	S98T002964	0	@IC-01 NO2-02	LIQUID	7.66e+04	7.78e+04	1.554 RPD
5	DUP	S98T002964	0	@IC-01 BR-02	LIQUID	<5.18e2	<5.18e2	RPD
5	DUP	S98T002964	0	@IC-01 NO3-02	LIQUID	1.74e+05	1.74e+05	0.000 RPD
5	DUP	S98T002964	0	@IC-01 PO4-02	LIQUID	9.23e+02	1.29e+03	33.168 RPD
5	DUP	S98T002964	0	@IC-01 SO4-02	LIQUID	1.69e+03	1.83e+03	7.955 RPD
5	DUP	S98T002964	0	@IC-01 OXALATE2	LIQUID	<4.35e2	<4.35e2	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26988

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 26988

_____
Analyst Signature   Date

_____
Analyst Signature   Date

11/9/98
_____
Reviewer Signature   Date

11/04/98 15:57
A-0004-1

Page: 1

LABCORP Data Entry Template for Worklist# 26988

Analyst: ADD Instrument: IC 4051 Book# 29N21-A ICV
 Method: LA-533-105 Rev/Mod F-0 28N21-A CCV
 Worklist Comment: BY105 FOR @IC-01 RTS

S	Type	Sample#	R	Test	Matrix	Group#	Project
1	CCB			@IC-QC	QC		
2	ICV			@IC-QC	QC		
3	CCV			@IC-QC	QC		
4	SAMPLE	S98T002964 0		@IC-01	LIQUID	98000509	BY-105 (2)
	Analytes Requested:			BR-02	CL-02	F-02	NO2-02 , NO3-02 ,
				OXALATE2,	PO4-02	SO4-02	
5	DUP	S98T002964 0		@IC-01	LIQUID		

Final page for worklist # 26988

Anthony Lurich 11-5-98
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 11-6-98
John Worell
26988NOV.CSV

validated 11/9/98 gmdave

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1245

Sample Name: 29N21-A ICV Date: 11/05/1998 14:59:05
 Data File : C:\DX\DATA\98110521.D09
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: A System: 1 Inject#: 9 Detector: CDM-1
 Analyst : *Anthony* Column: AG4A/AS4A anion column

41-5-98

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	93	360	2	
2	1.00	fluoride	59.376/00.64	2093	10211	2	1.01
3	1.47	chloride	88.222/10.28	2525	10932	1	-1.57
4	1.73	nitrite	565.686/09.29	9764	47612	1	-3.17
5	2.53	bromide	571.170/07.47	6240	34594	1	-0.26
6	2.86	nitrate	593.452/00.25	7363	48284	1	-2.05
7	4.21	phosphate	548.762/00.32	1609	19587	1	0.16
8	5.49	sulfate	654.011/02.51	5158	63789	1	0.06
9	7.15	oxalate	552.727/01.60	2515	40393	1	-0.05
Totals			3633.406	37361	275762		

File: 98110521.D09 Sample: 29N21-A ICV

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1246 TO 1250.

1246

```

=====
Sample Name: 28N21-A CCV                               Date: 11/05/1998 15:32:08
Data File  : C:\DX\DATA\98110521.D12
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 12                     Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101      3000 5Hz      0.00 10.00      30
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	89	355	2	
2	1.00	fluoride	60.758	2097	10451	2	1.01
3	1.47	chloride	93.460	2695	11600	1	-1.12
4	1.74	nitrite	550.246	9053	46286	1	-2.79
5	2.54	bromide	610.003	6548	37018	1	0.00
6	2.87	nitrate	716.415	8590	58622	1	-1.83
7	4.22	phosphate	648.026	1836	23251	1	0.48
8	5.49	sulfate	693.079	5180	67702	1	0.06
9	7.15	oxalate	559.694	2409	40908	1	-0.05
Totals			3931.681	38498	296192		

File: 98110521.D12 Sample: 28N21-A CCV

Sample Name: BLANK Date: 11/05/1998 14:46:19
Data File : C:\DX\DATA\98110521.D08
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 8 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.39	chloride	0.013	18	52	1	-6.94
2	2.92	nitrate	0.103	16	79	1	0.00
3	9.07		0.000	14	341	1	
Totals			0.115	48	472		

File: 98110521.D08 Sample: BLANK

1248

Data Reprocessed On 11/06/1998 09:26:54

```

=====
Sample Name: S98T002964 SAM      Date: 11/05/1998 15:56:24
Data File  : F:\DATA\98110521.D14
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 14      Detector: CDM-1
Analyst    :      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      4141    3000    5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	552	3128	2	
2	1.10		0.000	568	5699	2	
3	1.45	chloride	5223.537	3229	15957	1	-2.46
4	1.74	nitrite	76634.871	29577	162049	1	-2.79
5	2.79	nitrate	174011.886	47408	387639	1	-0.12
6	3.80		0.000	44	371	1	
7	4.24	phosphate	922.786	47	591	1	0.95
8	5.55	sulfate	1691.052	283	3518	1	1.03
Totals			258484.132	81708	578952		

File: 98110521.D14 Sample: S98T002964 SAM

Data Reprocessed On 11/06/1998 09:28:12

```

=====
Sample Name: S98T002964 DUP                      Date: 11/05/1998 16:07:49
Data File  : F:\DATA\98110521.D15
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 15              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          4141    3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	526	2973	2	
2	1.10		0.000	560	5452	2	
3	1.45	chloride	5333.041	3215	16302	1	-2.91
4	1.73	nitrite	77770.213	29268	164536	1	-3.17
5	2.54	bromide	400.775	15	65	1	0.00
6	2.78	nitrate	173639.718	47155	386693	1	0.00
7	3.81		0.000	31	226	1	
8	4.21	phosphate	1293.850	69	911	1	0.32
9	5.49	sulfate	1834.576	236	3857	1	0.06
10	7.20	oxalate	269.916	23	388	1	0.70
Totals			260542.089	81100	581404		

File: 98110521.D15 Sample: S98T002964 DUP

LABCORE Completed Worklist Report for Worklist# 27021

Analyst: ehc

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-O

Worklist Comment: BY-105, @ic4g-01, Was WKL26763 use 606df. jmf

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	IC4G-QC F*4	QC	6.10e1	5.79e+01	94.918 % Recovery	
2 LCS-INST	0	IC4G-QC ACETATE2	QC	1.16e2	1.05e+02	90.517 % Recovery	
2 LCS-INST	0	IC4G-QC FORMATE2	QC	1.19e2	1.17e+02	98.319 % Recovery	
2 LCS-INST	0	IC4G-QC GLYCOLT1	QC	1.08e2	1.00e+02	92.593 % Recovery	
3 CCV	0	IC4G-QC F*4	QC	6.29e1	5.29e+01	84.102 % Recovery	
3 CCV	0	IC4G-QC ACETATE2	QC	1.20e2	1.08e+02	90.000 % Recovery	
3 CCV	0	IC4G-QC FORMATE2	QC	1.20e2	1.16e+02	96.667 % Recovery	
3 CCV	0	IC4G-QC GLYCOLT1	QC	1.03e2	9.30e+01	90.291 % Recovery	
4 SAMPLE	S98T002960	IC4G-01 ACETATE2 LIQUID	N/A		1.693e+03	35.150 ug/mL	
4 SAMPLE	S98T002960	IC4G-01 FORMATE2 LIQUID	N/A		1.152e+03	37.570 ug/mL	
4 SAMPLE	S98T002960	IC4G-01 GLYCOLT1 LIQUID	N/A		3.394e+01	33.940 ug/mL	
5 DUP	S98T002960	IC4G-01 F*4-01 LIQUID		3.90e+02	4.22e+02	7.882 RPD	
5 DUP	S98T002960	IC4G-01 ACETATE2 LIQUID		1.69e+03	1.47e+03	13.924 RPD	
5 DUP	S98T002960	IC4G-01 FORMATE2 LIQUID		1.15e+03	1.14e+03	0.873 RPD	
5 DUP	S98T002960	IC4G-01 GLYCOLT1 LIQUID		<3.39e1	<3.39e1	RPD	
6 SPK	S98T002960	IC4G-01 F*4-01 LIQUID		6.10e1	5.49e+01	90.000 % Recovery	
6 SPK	S98T002960	IC4G-01 ACETATE2 LIQUID		1.16e2	7.91e+01	68.190 % Recovery	
6 SPK	S98T002960	IC4G-01 FORMATE2 LIQUID		1.19e2	1.32e+02	110.924 % Recovery	
6 SPK	S98T002960	IC4G-01 GLYCOLT1 LIQUID		1.08e2	1.14e+02	105.556 % Recovery	
7 SAMPLE	S98T002961	IC4G-01 ACETATE2 LIQUID	N/A		9.347e+02	35.150 ug/mL	
7 SAMPLE	S98T002961	IC4G-01 FORMATE2 LIQUID	N/A		6.436e+02	37.570 ug/mL	
7 SAMPLE	S98T002961	IC4G-01 GLYCOLT1 LIQUID	N/A		3.394e+01	33.940 ug/mL	
8 DUP	S98T002961	IC4G-01 F*4-01 LIQUID		1.60e+02	1.61e+02	0.623 RPD	
8 DUP	S98T002961	IC4G-01 ACETATE2 LIQUID		9.35e+02	9.41e+02	0.640 RPD	
8 DUP	S98T002961	IC4G-01 FORMATE2 LIQUID		6.44e+02	6.52e+02	1.235 RPD	
8 DUP	S98T002961	IC4G-01 GLYCOLT1 LIQUID		<3.39e1	<3.39e1	RPD	

Final page for worklist# 27021

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John M. Lyle
Reviewer Signature _____ Date 2/25/99

S98T002960 + S98T002961 reanalyzed for F only on W/L# 28395.
John 3/23/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

02/08/99 10:40

ws2

LABCORE Data Entry Template for Worklist# 27021

Analyst: P.H.C. Instrument: ICBook# 59N21-AMethod: LA-533-105 Rev/Mod F-058N21-A

Worklist Comment: BY-105, @ic4g-01, Was WKL26763 use 606df. jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T002960 0		@IC4G-01 LIQUID		98000509 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
5 DUP	S98T002960 0		@IC4G-01 LIQUID			
6 SPK	S98T002960 0		@IC4G-01 LIQUID			
7 SAMPLE	S98T002961 0		@IC4G-01 LIQUID		98000509 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
8 DUP	S98T002961 0		@IC4G-01 LIQUID			

Final page for worklist # 27021

P.H.C. 2-23-99
 Signature Date

_____
 Signature Date

Data Entry Comments:

uploaded 2-23-99J. Wainall27021FEB.CSV

registered jmf
5/11/00
Exempted on 598T002960 + 2961
Rest validated 2/25/99 JMF

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1252

=====

Sample Name: INSTR BLANK Date: 02/23/1999 08:25:44

Data File : C:\DX\DATA\99022311.D10

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 10 Detector: CDM-2

Analyst : *W. Cohen* Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	4954	5Hz	0.00	16.51	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	87803	1534980	1	
2	1.00		0.000	59831	930858	1	
3	2.43		0.000	213971	31795251	2	
4	4.88		0.000	1132267	46777741	2	
5	5.81		0.000	1710731	76210954	2	
6	5.97		0.000	1850377	14648844	2	
7	6.04		0.000	1881095	31427513	2	
Totals			0.000	6936075	203326141		

File: 99022311.D10 Sample: INSTR BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1253 TO 1262

W. Cohen

Data Reprocessed On 02/23/1999 13:46:39

Sample Name: 59N21-A LCS	Date: 02/23/1999 09:00:17
Data File : E:\DATA\99022311.D12	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 12	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4440	5Hz	0.00	14.80		16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	5647	31390	2	
2	0.40		0.000	2679	32357	2	
3	2.10		0.000	16897	310923	2	
4	2.29		0.000	11027	97287	3	
7	2.64		0.000	1379	35068	1	
11	3.43		0.000	2208	58345	2	
12	3.76		0.000	4551	36365	2	
13	3.95		0.000	3442	28513	2	
14	4.09		0.000	4694	42844	2	
15	4.37		0.000	1891	20109	2	
17	4.99		0.000	5189	76925	2	
18	5.23		0.000	4333	20775	2	
19	5.35		0.000	5016	61767	2	
20	5.73		0.000	4609	22853	2	
21	5.81		0.000	5194	75137	2	
22	6.23		0.000	2645	23308	1	
24	7.17	fluoride	57.919	1774509	26333234	2	0.00
25	7.87	glycolate	100.199	489680	9078708	2	0.00
26	8.35	acetate	104.509	496607	9541138	2	0.00
27	9.04	formate	117.260	919478	17976052	2	0.00
Totals			379.887	3761676	63903098		

Sample Name: 58N21-A CCV

Date: 02/23/1999 09:24:40

Data File : C:\DX\DATA\99022311.D01

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 1

Detector: CDM-2

Analyst : *EM* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 4320 5Hz 0.00 14.40 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	17180	428612	1	
6	2.80		0.000	4156	71854	3	
9	3.73		0.000	3075	23727	2	
10	3.99		0.000	3634	47567	2	
11	4.33		0.000	19649	160727	2	
12	4.51		0.000	22087	406245	2	
16	5.84		0.000	900	29702	2	
17	5.99		0.000	425	25274	2	
20	7.23	fluoride	52.891	1659734	23928740	2	0.00
21	7.89	glycolate	92.987	453790	8398984	2	0.00
22	8.37	acetate	108.273	514407	9891098	2	0.00
23	9.07	formate	116.169	932697	17800035	2	0.00
Totals			370.321	3631734	61212565		

File: 99022311.D01 Sample: 58N21-A CCV

Sample Name: S98T002960

Date: 02/23/1999 10:22:08

Data File : C:\DX\DATA\99022311.D02

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 2

Detector: CDM-2

Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	606	4571	5Hz	0.00	15.23	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	34186	811272	1	
2	1.70		0.000	194606	2420660	1	
3	2.71		0.000	9245	31856	2	
4	2.92		0.000	21531	557938	2	
5	3.33		0.000	23079	847627	2	
6	3.92		0.000	21341	182086	2	
7	4.11		0.000	22198	384988	2	
8	4.40		0.000	19609	129817	2	
9	4.56		0.000	20659	182324	2	
10	4.71		0.000	21064	342500	2	
11	4.97		0.000	19867	284172	2	
12	5.24		0.000	19030	166490	2	
13	5.44		0.000	18912	732433	2	
14	7.09	fluoride	390.384	1705896	29758411	2	0.00
15	8.27	acetate	1693.050	1078812	25886451	2	0.00
16	8.96	formate	1151.797	1556262	30150935	2	0.00
17	9.73		0.000	247857	4836521	2	

Totals	3235.231	5034154	97706481

File: 99022311.D02 Sample: S98T002960

```

=====
Sample Name: S98T002960 DUP          Date: 02/23/1999 11:54:05
Data File  : C:\DX\DATA\99022311.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3
Analyst    :          Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606    4499  5Hz    0.00  14.99    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	22402	529904	1	
2	1.70		0.000	225419	2731292	1	
3	3.13		0.000	3504	108704	3	
5	3.40		0.000	2080	18966	2	
7	3.93		0.000	1346	20756	1	
10	4.75		0.000	3271	28184	2	
11	4.89		0.000	3630	53057	2	
12	5.72		0.000	25710	381902	2	
13	5.81		0.000	26734	625041	2	
14	7.12	fluoride	422.162	1826632	32303665	2	0.00
15	8.27	acetate	1470.183	1046197	22497820	2	0.00
16	8.96	formate	1144.102	1581401	29939836	2	0.00
17	9.71		0.000	246260	4764562	2	
Totals			3036.447	5014585	94003689		

File: 99022311.D03 Sample: S98T002960 DUP

Data Reprocessed On 02/23/1999 13:36:34

Sample Name: S98T002960 SPK	Date: 02/23/1999 12:11:06
Data File : E:\DATA\99022311.D04	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 4	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	606	4765	5Hz	0.00	15.88	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	3834	37342	1	
2	1.70		0.000	165507	1935468	1	
3	2.92		0.000	25253	387024	2	
4	3.17		0.000	27827	626121	2	
5	3.40		0.000	26557	694410	2	
6	3.91		0.000	24026	541565	2	
7	4.31		0.000	23296	663132	2	
8	4.76		0.000	22695	147625	2	
9	4.88		0.000	23426	97485	2	
10	4.96		0.000	23500	378490	2	
11	5.27		0.000	24284	1040338	2	
12	7.01	fluoride	664.812	3187306	51916562	2	0.00
13	7.71	glycolate	571.245	469818	8608005	2	0.00
14	8.19	acetate	2088.438	1477912	31862017	2	0.00
15	8.88	formate	1812.001	2479724	48544719	2	0.00
16	9.65		0.000	247569	4914941	2	
Totals			5136.496	8252536	152395245		

File: 99022311.D04 Sample: S98T002960 SPK

```

=====
Sample Name: S98T002961      Date: 02/23/1999 12:27:31
Data File   : C:\DX\DATA\99022311.D05
Method      : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5
Analyst     :      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606    4620  5Hz    0.00  15.40      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	5268	90670	1	
2	1.73		0.000	78409	813552	1	
3	2.92		0.000	13819	241035	2	
4	2.99		0.000	13064	266513	2	
5	3.49		0.000	4526	54539	2	
10	4.11		0.000	853	59307	2	
11	4.79		0.000	6475	207400	2	
12	5.24		0.000	9789	63090	2	
13	5.40		0.000	11581	125020	2	
14	5.53		0.000	11971	430592	2	
15	7.09	fluoride	160.190	705354	11477247	2	0.00
16	8.27	acetate	934.741	653409	14295011	2	0.00
17	8.96	formate	643.626	853714	16366258	2	0.00
18	9.76		0.000	142697	2774513	2	
Totals			1738.556	2510929	47264748		

File: 99022311.D05 Sample: S98T002961

```

=====
Sample Name: S98T002961 DUP                      Date: 02/23/1999 12:42:47
Data File  : C:\DX\DATA\99022311.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606    4256  5Hz   0.00  14.18      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	3531	44248	1	
2	1.73		0.000	83717	888848	1	
3	2.87		0.000	18531	383255	2	
4	3.05		0.000	18296	454387	2	
5	3.48		0.000	17249	472886	2	
6	3.97		0.000	16213	209128	2	
7	4.19		0.000	16579	91240	2	
8	4.31		0.000	16193	196089	2	
9	4.52		0.000	16719	121690	2	
10	4.68		0.000	16804	108406	2	
11	4.79		0.000	17047	108276	2	
12	4.89		0.000	17430	114366	2	
13	5.01		0.000	17043	71409	2	
14	5.09		0.000	17226	70002	2	
15	5.17		0.000	17568	192810	2	
16	5.40		0.000	17265	194091	2	
17	5.59		0.000	17750	511021	2	
18	7.09	fluoride	161.428	736254	11574911	2	0.00
19	8.24	acetate	940.934	658645	14390398	2	0.00
20	8.96	formate	652.115	879239	16594015	2	0.00
21	9.73		0.000	141433	2735881	2	
Totals			1754.478	2760732	49527356		

LABCORE Completed Worklist Report for Worklist# 27024

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-D

Worklist Comment: BY-105, @IC4G-01, Was 26764. Use 606DF. jmf

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL	
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL	
1 CCB	0	@IC4G-QC FORMATE2	QC	1	4.30e-02	0.043	ug/mL	
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	8.20e-02	0.082	ug/mL	
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	6.30e+01	103.279 %	Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.15e+02	99.138 %	Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.24e+02	104.202 %	Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.08e2	1.06e+02	98.148 %	Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.29e1	6.01e+01	95.548 %	Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.22e+02	101.667 %	Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.27e+02	105.833 %	Recovery	
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.03e2	1.02e+02	99.029 %	Recovery	
5 DUP	S98T002962	0	@IC4G-01 F*4-01	LIQUID	1.10e+02	1.81e+02	48.797 RPD	
5 DUP	S98T002962	0	@IC4G-01 ACETATE2	LIQUID	9.06e+02	2.01e+03	75.720 RPD	
5 DUP	S98T002962	0	@IC4G-01 FORMATE2	LIQUID	6.42e+02	1.33e+03	69.777 RPD	
5 DUP	S98T002962	0	@IC4G-01 GLYCOLT1	LIQUID	4.32e+01	<3.39e1	RPD	
6 SPK	S98T002962	0	@IC4G-01 F*4-01	LIQUID	6.10e1	7.29e+01	119.508 %	Recovery
6 SPK	S98T002962	0	@IC4G-01 ACETATE2	LIQUID	1.16e2	3.32e+02	286.207 %	Recovery
6 SPK	S98T002962	0	@IC4G-01 FORMATE2	LIQUID	1.19e2	2.83e+02	237.815 %	Recovery
6 SPK	S98T002962	0	@IC4G-01 GLYCOLT1	LIQUID	1.08e2	1.20e+02	111.111 %	Recovery
7 SAMPLE	S98T002963	0	@IC4G-01 F*4-01	LIQUID	N/A	1.001e+02	18.790 ug/mL	
7 SAMPLE	S98T002963	0	@IC4G-01 ACETATE2	LIQUID	N/A	2.255e+03	35.150 ug/mL	
7 SAMPLE	S98T002963	0	@IC4G-01 FORMATE2	LIQUID	N/A	1.425e+03	37.570 ug/mL	
7 SAMPLE	S98T002963	0	@IC4G-01 GLYCOLT1	LIQUID	N/A	<3.39e1	33.940 ug/mL	
8 DUP	S98T002963	0	@IC4G-01 F*4-01	LIQUID	1.00e+02	1.02e+02	1.980 RPD	
8 DUP	S98T002963	0	@IC4G-01 ACETATE2	LIQUID	2.26e+03	2.26e+03	0.000 RPD	
8 DUP	S98T002963	0	@IC4G-01 FORMATE2	LIQUID	1.42e+03	1.42e+03	0.000 RPD	
8 DUP	S98T002963	0	@IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD	
9 SPK	S98T002963	0	@IC4G-01 F*4-01	LIQUID	6.10e1	5.07e+01	83.115 %	Recovery
9 SPK	S98T002963	0	@IC4G-01 ACETATE2	LIQUID	1.16e2	1.06e+02	91.379 %	Recovery
9 SPK	S98T002963	0	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.46e+02	122.689 %	Recovery
9 SPK	S98T002963	0	@IC4G-01 GLYCOLT1	LIQUID	1.08e2	1.32e+02	122.222 %	Recovery

Final page for worklist# 27024

Analyst Signature

Date

Analyst Signature

Date

Grand Lays 2/25/99

3/23/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27024

Analyst: F-000 Instrument: IC 4552 Book# 59N21-A
Method: LA-533-105 Rev/Mod F-0 59N21-A LCS
58N21-N CCV
Worklist Comment: BY-105, @IC4G-01, Was 26764. Use 606DF. jmf

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC QC			
2	LCS-INST				@IC4G-QC QC			
3	CCV				@IC4G-QC QC			
4	SAMPLE	S98T002962 0			@IC4G-01 LIQUID		98000509	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
5	DUP	S98T002962 0			@IC4G-01 LIQUID			
6	SPK	S98T002962 0			@IC4G-01 LIQUID			
7	SAMPLE	S98T002963 0			@IC4G-01 LIQUID		98000509	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
8	DUP	S98T002963 0			@IC4G-01 LIQUID			
9	SPK	S98T002963 0			@IC4G-01 LIQUID			

Final page for worklist # 27024

Anthony P. ... 02-22-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 2-23-99

Sh. Howell

27024 FEB. CSV

598T002962 rejected for

high RPD's & poor spk Rest

Validated 2/25/99 gmf

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1264

Sample Name: *ICV 59N21-A* *Jan 2-28-98* Date: 02/22/1999 14:15:47
 Data File : C:\DX\DATA\99022011.D02
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2
 Analyst : *AMF for ADP* Column: AG14A-SC, AS14A-SC, SRS
3/25/99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	21852	663422	2	
3	2.13		0.000	13527	321767	2	
4	2.28		0.000	3702	22976	2	
7	2.76		0.000	4364	26152	2	
8	2.92		0.000	5358	44588	2	
9	3.19		0.000	6613	80597	2	
11	3.49		0.000	3622	54766	2	
12	3.92		0.000	3431	45819	2	
13	4.17		0.000	3789	58231	2	
18	5.05		0.000	2273	18901	2	
20	5.31		0.000	3119	21608	2	
24	5.95		0.000	8767	112201	2	
25	6.17		0.000	9924	203788	2	
26	7.31	fluoride	63.036/03.3	1880894	28785129	2	0.00
27	8.03	glycolate	106.12098.2	527181	9638481	2	0.00
28	8.51	acetate	114.54598.7	527172	10473918	2	0.00
29	9.23	formate	124.336/04.4	1004059	19118802	2	0.00

Totals 408.037 4029649 69691145

File: 99022011.D02 Sample: ICV 59N21-A

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1265 TO 1272.

Sample Name: CCV 58N21-A Date: 02/22/1999 14:38:05
 Data File : C:\DX\DATA\99022011.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	28332	1410880	1	
2	2.29		0.000	3504	32401	2	
4	2.60		0.000	3987	34415	2	
5	2.77		0.000	4393	58533	2	
14	4.71		0.000	4023	37852	2	
15	4.88		0.000	4343	21532	2	
16	5.01		0.000	4939	27020	2	
17	5.16		0.000	3036	26225	2	
21	5.83		0.000	4252	37902	1	
27	7.33	fluoride	60.073	1787442	27364898	2	0.00
28	8.05	glycolate	101.797	500848	9229645	2	0.00
29	8.56	acetate	121.677	571078	11135953	2	0.00
30	9.25	formate	127.436	1020525	19620123	2	0.00
Totals			410.983	3940702	69037381		

File: 99022011.D03 Sample: CCV 58N21-A

1266

Sample Name: CCB BLANK Date: 02/22/1999 13:50:04
 Data File : C:\DX\DATA\99022011.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	50089	1691584	1	
3	2.00		0.000	30897	267537	2	
4	2.27		0.000	34361	373545	2	
5	2.48		0.000	34595	266404	2	
6	2.67		0.000	32116	228580	2	
7	2.80		0.000	32827	211234	2	
8	2.89		0.000	32787	627560	2	
9	3.24		0.000	26753	172322	2	
10	3.36		0.000	25648	515901	2	
11	3.76		0.000	21023	165156	2	
12	3.92		0.000	19349	208366	2	
13	4.13		0.000	18563	104088	2	
14	4.21		0.000	18206	222940	2	
15	4.49		0.000	12647	161767	2	
16	4.72		0.000	11278	59347	2	
17	4.88		0.000	11824	206197	2	
18	5.20		0.000	6276	34382	2	
19	5.35		0.000	5705	47786	2	
27	6.33		0.000	2283	39786	1	
28	7.04		0.000	2891	32846	2	
29	7.65		0.000	20920	466026	2	
30	8.08	glycolate	0.082	24724	574431	2	0.00
31	8.32		0.000	26982	922618	3	
32	9.17	formate	0.043	1855	17288	4	0.00
35	9.87		0.000	1884	19190	1	
Totals			0.125	506484	7636880		

File: 99022011.D01 Sample: CCB BLANK

1268

```

=====
Sample Name: S98T002963 SAM                      Date: 02/22/1999 16:24:23
Data File  : E:\DATA\99022011.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606  5700  5Hz  0.00 19.00      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	14534	444184	1	
3	1.70		0.000	198955	2370566	1	
4	3.11		0.000	30784	769405	2	
5	3.23		0.000	30048	167169	2	
6	3.36		0.000	30374	314208	2	
7	3.52		0.000	30886	178717	2	
8	3.73		0.000	34709	622287	2	
9	3.99		0.000	34980	553891	2	
10	4.33		0.000	34069	347415	2	
11	4.43		0.000	34567	505078	2	
12	4.79		0.000	38619	846913	2	
13	5.11		0.000	38483	422543	2	
14	5.32		0.000	40246	1023330	2	
15	5.72		0.000	37746	1076657	2	
16	7.17	fluoride	100.054	356446	6745710	2	0.00
17	8.37	acetate	2255.376	1449900	34371275	2	0.00
18	9.07	formate	1424.527	1889340	37680542	2	0.00
19	9.84		0.000	315078	6276162	2	
Totals			3779.956	4639763	94716051		

File: 99022011.D07 Sample: S98T002963 SAM

Data Reprocessed On 02/23/1999 11:07:53

```

=====
Sample Name: S98T002963  DUP                      Date: 02/22/1999 16:44:00
Data File   : E:\DATA\99022011.D08
Method      : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 8
Analyst     :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606   5700  5Hz   0.00  19.00      16000
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	4482	71142	2	
3	1.70		0.000	196051	2365982	1	
4	3.07		0.000	32632	645110	2	
5	3.13		0.000	33016	384617	2	
6	3.43		0.000	31468	278617	2	
7	3.53		0.000	32144	158968	2	
8	3.61		0.000	32285	160940	2	
9	3.73		0.000	34357	222614	2	
10	3.85		0.000	35169	1136506	2	
11	4.48		0.000	36944	492084	2	
12	4.61		0.000	37733	247769	2	
13	4.75		0.000	39162	253053	2	
14	4.87		0.000	38914	471942	2	
15	5.09		0.000	38782	193315	2	
16	5.17		0.000	38983	405307	2	
17	5.39		0.000	40354	989134	2	
18	5.80		0.000	36758	863580	2	
19	7.17	fluoride	101.650	368426	6871124	2	0.00
20	8.35	acetate	2257.532	1438662	34403628	2	0.00
21	9.07	formate	1417.058	1883060	37473072	2	0.00
22	9.84		0.000	308608	6153996	2	
Totals			3776.240	4737990	94242500		

```

=====
Sample Name: S98T002963  SPK                      Date: 02/22/1999 17:03:47
Data File  : E:\DATA\99022011.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           606   5700  5Hz   0.00  19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	8442	348236	1	
2	1.70		0.000	195890	2361412	1	
3	3.19		0.000	11887	533865	3	
7	4.07		0.000	3151	65216	2	
8	4.33		0.000	2590	22292	2	
14	5.31		0.000	1842	95357	2	
15	5.69		0.000	9573	302939	2	
16	7.17	fluoride	353.743	1624044	26830275	2	0.00
17	7.87	glycolate	658.254	552679	9978559	2	0.00
18	8.35	acetate	2787.276	1861349	42313222	2	0.00
19	9.04	formate	2152.939	2832485	58273829	2	0.00
20	9.84		0.000	311602	6198932	2	
Totals			5952.212	7415533	147324134		

File: 99022011.D09 Sample: S98T002963 SPK

LABCORE Completed Worklist Report for Worklist# 27276

Analyst: adp

Instrument: IC40S1

Book#: 29N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 (2), @IC-01 skm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC PO4	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC SO4	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	LCS-INST	0	@IC-QC F	QC	5.90e1	6.04e+01	102.373 % Recovery	
2	LCS-INST	0	@IC-QC CL	QC	8.00e1	8.21e+01	102.625 % Recovery	
2	LCS-INST	0	@IC-QC NO2	QC	5.48e2	5.61e+02	102.372 % Recovery	
2	LCS-INST	0	@IC-QC BR	QC	5.86e2	5.88e+02	100.341 % Recovery	
2	LCS-INST	0	@IC-QC NO3	QC	5.92e2	5.79e+02	97.804 % Recovery	
2	LCS-INST	0	@IC-QC PO4	QC	5.47e2	5.74e+02	104.936 % Recovery	
2	LCS-INST	0	@IC-QC SO4	QC	6.38e2	6.56e+02	102.821 % Recovery	
2	LCS-INST	0	@IC-QC OXALATE2	QC	5.40e2	5.74e+02	106.296 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.04e+01	94.375 % Recovery	
3	CCV	0	@IC-QC CL	QC	9.00e1	9.43e+01	104.778 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.62e2	5.52e+02	98.221 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.01e+02	95.397 % Recovery	
3	CCV	0	@IC-QC NO3	QC	6.98e2	6.88e+02	98.567 % Recovery	
3	CCV	0	@IC-QC PO4	QC	6.32e2	6.46e+02	102.215 % Recovery	
3	CCV	0	@IC-QC SO4	QC	6.99e2	7.01e+02	100.286 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.53e2	5.78e+02	104.521 % Recovery	
4	BLNK-PREP	0	@IC-01 F-02	SOLID	1	<1.20e-2		ug/g
4	BLNK-PREP	0	@IC-01 CL-02	SOLID	1	<1.70e-2		ug/g
4	BLNK-PREP	0	@IC-01 NO2-02	SOLID	1	5.25e-01	0.525	ug/g
4	BLNK-PREP	0	@IC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4	BLNK-PREP	0	@IC-01 NO3-02	SOLID	1	1.74e-01	0.174	ug/g
4	BLNK-PREP	0	@IC-01 PO4-02	SOLID	1	<1.20e-1		ug/g
4	BLNK-PREP	0	@IC-01 SO4-02	SOLID	1	<1.38e-1		ug/g
4	BLNK-PREP	0	@IC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5	SAMPLE	S98T002994	0 W @IC-01 F-02	SOLID	N/A	2.079e+03	119.500	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 CL-02	SOLID	N/A	1.103e+03	169.300	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 NO2-02	SOLID	N/A	1.361e+04	1076.000	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 BR-02	SOLID	N/A	1.245e+03	1245.000	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 PO4-02	SOLID	N/A	7.694e+03	1195.000	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 SO4-02	SOLID	N/A	5.783e+03	1374.000	ug/g
5	SAMPLE	S98T002994	0 W @IC-01 OXALATE2	SOLID	N/A	2.682e+03	1046.000	ug/g
6	DUP	S98T002994	0 W @IC-01 F-02	SOLID	2.08e+03	2.04e+03	1.942	RPD
6	DUP	S98T002994	0 W @IC-01 CL-02	SOLID	1.10e+03	1.04e+03	5.607	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27276

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002994	0 W	④IC-01	NO2-02	SOLID	1.36e+04	1.29e+04	5.283 RPD
6 DUP	S98T002994	0 W	④IC-01	BR-02	SOLID	<1.24e3	<1.25e3	RPD
6 DUP	S98T002994	0 W	④IC-01	NO3-02	SOLID	4.88e+05	4.81e+05	1.445 RPD
6 DUP	S98T002994	0 W	④IC-01	PO4-02	SOLID	7.69e+03	6.90e+03	10.829 RPD
6 DUP	S98T002994	0 W	④IC-01	SO4-02	SOLID	5.78e+03	5.65e+03	2.275 RPD
6 DUP	S98T002994	0 W	④IC-01	OXALATE2	SOLID	2.68e+03	2.82e+03	5.091 RPD
7 SPK	S98T002994	0 W	④IC-01	F-02	SOLID	5.90e1	5.93e+01	100.508 % Recovery
7 SPK	S98T002994	0 W	④IC-01	CL-02	SOLID	8.00e1	8.08e+01	101.000 % Recovery
7 SPK	S98T002994	0 W	④IC-01	NO2-02	SOLID	5.48e2	5.57e+02	101.642 % Recovery
7 SPK	S98T002994	0 W	④IC-01	BR-02	SOLID	5.86e2	5.43e+02	92.662 % Recovery
7 SPK	S98T002994	0 W	④IC-01	NO3-02	SOLID	5.92e2	3.39e+01	5.726 % Recovery
7 SPK	S98T002994	0 W	④IC-01	PO4-02	SOLID	5.47e2	5.48e+02	100.183 % Recovery
7 SPK	S98T002994	0 W	④IC-01	SO4-02	SOLID	6.38e2	6.34e+02	99.373 % Recovery
7 SPK	S98T002994	0 W	④IC-01	OXALATE2	SOLID	5.40e2	5.44e+02	100.741 % Recovery
8 SAMPLE	S98T002995	0 W	④IC-01	CL-02	SOLID	N/A	1.312e+03	171.000 ug/g
8 SAMPLE	S98T002995	0 W	④IC-01	NO2-02	SOLID	N/A	1.791e+04	1086.000 ug/g
8 SAMPLE	S98T002995	0 W	④IC-01	BR-02	SOLID	N/A	1.257e+03	1257.000 ug/g
9 DUP	S98T002995	0 W	④IC-01	F-02	SOLID	2.93e+03	6.92e+03	81.015 RPD
9 DUP	S98T002995	0 W	④IC-01	CL-02	SOLID	1.31e+03	1.26e+03	3.891 RPD
9 DUP	S98T002995	0 W	④IC-01	NO2-02	SOLID	1.79e+04	1.65e+04	8.140 RPD
9 DUP	S98T002995	0 W	④IC-01	BR-02	SOLID	<1.26e3	<1.26e3	RPD
9 DUP	S98T002995	0 W	④IC-01	NO3-02	SOLID	3.62e+05	1.33e+05	92.525 RPD
9 DUP	S98T002995	0 W	④IC-01	PO4-02	SOLID	8.68e+03	3.10e+03	94.737 RPD
9 DUP	S98T002995	0 W	④IC-01	SO4-02	SOLID	7.26e+03	3.28e+03	75.522 RPD
9 DUP	S98T002995	0 W	④IC-01	OXALATE2	SOLID	6.74e+03	1.19e+04	55.365 RPD
10 SPK	S98T002995	0 W	④IC-01	F-02	SOLID	5.90e1	6.07e+01	102.881 % Recovery
10 SPK	S98T002995	0 W	④IC-01	CL-02	SOLID	8.00e1	8.55e+01	106.875 % Recovery
10 SPK	S98T002995	0 W	④IC-01	NO2-02	SOLID	5.48e2	5.91e+02	107.847 % Recovery
10 SPK	S98T002995	0 W	④IC-01	BR-02	SOLID	5.86e2	5.56e+02	96.587 % Recovery
10 SPK	S98T002995	0 W	④IC-01	NO3-02	SOLID	5.92e2	9.06e+02	153.041 % Recovery
10 SPK	S98T002995	0 W	④IC-01	PO4-02	SOLID	5.47e2	5.64e+02	103.108 % Recovery
10 SPK	S98T002995	0 W	④IC-01	SO4-02	SOLID	6.38e2	6.76e+02	105.956 % Recovery
10 SPK	S98T002995	0 W	④IC-01	OXALATE2	SOLID	5.40e2	5.69e+02	105.370 % Recovery
11 SAMPLE	S98T002996	0 W	④IC-01	CL-02	SOLID	N/A	6.593e+02	171.700 ug/g
11 SAMPLE	S98T002996	0 W	④IC-01	NO2-02	SOLID	N/A	7.852e+03	1030.000 ug/g
11 SAMPLE	S98T002996	0 W	④IC-01	BR-02	SOLID	N/A	1.262e+03	1262.000 ug/g
11 SAMPLE	S98T002996	0 W	④IC-01	NO3-02	SOLID	N/A	4.454e+05	1404.000 ug/g
11 SAMPLE	S98T002996	0 W	④IC-01	PO4-02	SOLID	N/A	8.688e+03	1212.000 ug/g
11 SAMPLE	S98T002996	0 W	④IC-01	SO4-02	SOLID	N/A	5.217e+03	1393.000 ug/g
12 DUP	S98T002996	0 W	④IC-01	F-02	SOLID	2.24e+03	2.97e+03	28.023 RPD
12 DUP	S98T002996	0 W	④IC-01	CL-02	SOLID	6.59e+02	6.64e+02	0.756 RPD
12 DUP	S98T002996	0 W	④IC-01	NO2-02	SOLID	7.85e+03	7.63e+03	2.842 RPD
12 DUP	S98T002996	0 W	④IC-01	BR-02	SOLID	<1.26e3	<1.25e3	RPD
12 DUP	S98T002996	0 W	④IC-01	NO3-02	SOLID	4.45e+05	3.94e+05	12.157 RPD
12 DUP	S98T002996	0 W	④IC-01	PO4-02	SOLID	8.69e+03	1.02e+04	15.987 RPD
12 DUP	S98T002996	0 W	④IC-01	SO4-02	SOLID	5.22e+03	6.04e+03	14.565 RPD
12 DUP	S98T002996	0 W	④IC-01	OXALATE2	SOLID	7.25e+03	5.65e+03	24.806 RPD
13 SAMPLE	S98T002997	0 W	④IC-01	F-02	SOLID	N/A	1.621e+03	120.500 ug/g
13 SAMPLE	S98T002997	0 W	④IC-01	CL-02	SOLID	N/A	1.203e+03	170.600 ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

1274

LABCORE Completed Worklist Report for Worklist# 27276

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
13 SAMPLE	S98T002997	0 W	@IC-01	NO2-02	SOLID	N/A	1.319e+04	1084.000 ug/g
13 SAMPLE	S98T002997	0 W	@IC-01	BR-02	SOLID	N/A	1.255e+03	1255.000 ug/g
13 SAMPLE	S98T002997	0 W	@IC-01	PO4-02	SOLID	N/A	6.909e+03	1205.000 ug/g
13 SAMPLE	S98T002997	0 W	@IC-01	SO4-02	SOLID	N/A	5.039e+03	1386.000 ug/g
13 SAMPLE	S98T002997	0 W	@IC-01	OXALATE	SOLID	N/A	2.619e+03	1054.000 ug/g
14 DUP	S98T002997	0 W	@IC-01	F-02	SOLID	1.62e+03	1.60e+03	1.242 RPD
14 DUP	S98T002997	0 W	@IC-01	CL-02	SOLID	1.20e+03	9.85e+02	19.680 RPD
14 DUP	S98T002997	0 W	@IC-01	NO2-02	SOLID	1.32e+04	1.22e+04	7.874 RPD
14 DUP	S98T002997	0 W	@IC-01	BR-02	SOLID	<1.26e3	<1.26e3	RPD
14 DUP	S98T002997	0 W	@IC-01	NO3-02	SOLID	5.04e+05	5.25e+05	4.082 RPD
14 DUP	S98T002997	0 W	@IC-01	PO4-02	SOLID	6.91e+03	5.80e+03	17.467 RPD
14 DUP	S98T002997	0 W	@IC-01	SO4-02	SOLID	5.04e+03	5.97e+03	16.894 RPD
14 DUP	S98T002997	0 W	@IC-01	OXALATE	SOLID	2.62e+03	2.50e+03	4.688 RPD

Final page for worklist# 27276

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jeanne Faye
Reviewer Signature _____ Date *11/23/98*

S98T002994 reanalyzed for NO₃ only on WL# 27291.
S98T002995 reanalyzed for F, NO₃, PO₄ & SO₄ + oxalate only
on WL# 27292.
S98T002996 reanalyzed for F & oxalate on WL# 27291.
S98T002997 reanalyzed for NO₃ only on WL# 27292.
J. Faye 3/23/99

LABCORE Data Entry Template for Worklist# 27276

Analyst: ADP Instrument: IC 4051 Book# 29N21-B LCS
Method: LA-533-105 Rev/Mod F-0 28N21-B CCV
Worklist Comment: BY105 (2), @IC-01 skm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002994 0 W		@IC-01	SOLID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
6 DUP	S98T002994 0 W		@IC-01	SOLID		
7 SPK	S98T002994 0 W		@IC-01	SOLID		
8 SAMPLE	S98T002995 0 W		@IC-01	SOLID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
9 DUP	S98T002995 0 W		@IC-01	SOLID		
10 SPK	S98T002995 0 W		@IC-01	SOLID		
11 SAMPLE	S98T002996 0 W		@IC-01	SOLID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
12 DUP	S98T002996 0 W		@IC-01	SOLID		
13 SAMPLE	S98T002997 0 W		@IC-01	SOLID	98000509	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
14 DUP	S98T002997 0 W		@IC-01	SOLID		

Data Entry Comments

uploaded 11-23-98
John W. Wall

The following were rejected:

S98T002994 - NO₃

S98T002995 - F, NO₃, PO₄, SO₄, Oxalate

S98T002996 - F, Oxalate

S98T002997 - NO₃

The rest was validated. JMW
11/23/98

1276

LABCORE Data Entry Template for Worklist# 27276

S Type	Sample#	R A	Test	Matrix	Group#	Project

Final page for worklist # 27276

Anthony Parente

11-22-98

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1277

Sample Name: 29N21-B LCS Date: 11/22/1998 10:02:58
 Data File : C:\DX\DATA\98112201.D14
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 14 Detector: CDM-1
 Analyst : *Anthony Furuta* Column: AG4A/AS4A anion column

11-22-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	75	266	2	
2	1.01	fluoride	60.444	102.45	2004	10396	2 2.36
3	1.49	chloride	82.137	102.67	2303	10159	1 0.22
4	1.77	nitrite	560.677	102.31	9073	47182	1 -0.93
5	2.63	bromide	588.211	100.38	5932	35657	1 3.67
6	2.97	nitrate	578.891	97.79	6806	47064	1 1.83
7	4.17	phosphate	573.587	104.86	1607	20501	1 -0.63
8	5.49	sulfate	656.156	102.95	5334	64004	1 0.06
9	7.15	oxalate	573.987	106.29	2556	41966	1 -0.05
Totals			3674.090	35688	277195		

File: 98112201.D14 Sample: 29N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1278 TO 1291

1278

Sample Name: 28N21-B CCV Date: 11/22/1998 09:50:57
 Data File : C:\DX\DATA\98112201.D13
 Method : C:\DX\METHOD\KIT.MBT
 ACI Address: 1 System: 1 Inject#: 13 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	88	328	2	
2	0.99	fluoride	60.436	2001	10395	2	0.34
3	1.47	chloride	94.272	2659	11703	1	-1.12
4	1.75	nitrite	552.230	9095	46456	1	-2.05
5	2.61	bromide	600.884	6278	36448	1	2.62
6	2.93	nitrate	688.137	8115	56239	1	0.46
7	4.14	phosphate	645.942	1847	23173	1	-1.43
8	5.44	sulfate	700.898	5625	68486	1	-0.91
9	7.09	oxalate	578.394	2567	42293	1	-0.79

Totals 3921.194 38276 295521

File: 98112201.D13 Sample: 28N21-B CCV

1279

Sample Name: BLANK Date: 11/22/1998 09:27:38
Data File : C:\DX\DATA\98112201.D11
Method : C:\DX\METHOD\KIT.MBT
ACI Address: 1 System: 1 Inject#: 11 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
Totals			0.000	0	0		

File: 98112201.D1 Sample: BLANK

Sample Name: PREP BLANK Date: 11/22/1998 10:13:54
 Data File : C:\DX\DATA\98112201.D15
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 15 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.28		0.000	61	567	1	
2	1.77	nitrite	0.525	788	3921	1	-0.93
3	3.01	nitrate	0.174	106	668	1	3.20
4	7.47	oxalate	0.073	25	449	1	4.43
Totals			0.773	980	5605		

File: 98112201.D15 Sample: PREP BLANK

1281

=====

Sample Name: S98T002994 SAM	Date: 11/22/1998 10:27:31
Data File : C:\DX\DATA\98112201.D16	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 16	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	117	501	2	
2	1.00	fluoride	10.646	606	3601	2	1.01
3	1.48	chloride	5.648	271	1273	1	-0.67
4	1.76	nitrite	69.696	2061	11115	1	-1.68
5	2.85	nitrate	2497.507	52102	462895	1	-2.51
6	4.17	phosphate	39.399	193	2562	1	-0.63
7	5.49	sulfate	29.617	417	5203	1	0.06
8	7.15	oxalate	13.733	106	1884	1	-0.05

Totals	2666.247	55872	489034

File: 98112201.D16 Sample: S98T002994 SAM

1282

Sample Name: S98T002994 DUP Date: 11/22/1998 10:44:07
 Data File : C:\DX\DATA\98112201.D17
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 17 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	132	561	2	
2	1.01	fluoride	10.362	587	3504	2	1.68
3	1.49	chloride	5.292	252	1186	1	-0.22
4	1.77	nitrite	65.816	1992	10463	1	-1.30
5	2.86	nitrate	2448.398	51038	452212	1	-2.05
6	4.19	phosphate	35.140	181	2262	1	-0.32
7	5.49	sulfate	28.777	386	5042	1	0.06
8	7.15	oxalate	14.358	99	1974	1	-0.05
Totals			2608.142	54667	477205		

File: 98112201.D17 Sample: S98T002994 DUP

1283

Sample Name: S98T002994 SPK Date: 11/22/1998 10:55:11
 Data File : C:\DX\DATA\98112201.D18
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	196	833	2	
2	1.01	fluoride	40.318	2482	13791	2	1.68
3	1.49	chloride	46.059	2401	11313	1	-0.22
4	1.77	nitrite	348.435	10995	58318	1	-1.30
5	2.62	bromide	271.421	6064	32497	1	3.15
6	2.85	nitrate	2514.437	52241	466599	1	-2.28
7	4.17	phosphate	313.559	1729	22249	1	-0.63
8	5.44	sulfate	346.545	4673	67022	1	-0.91
9	7.09	oxalate	285.486	2380	41329	1	-0.79
Totals			4166.259	83159	713951		

File: 98112201.D18 Sample: S98T002994 SPK

1284

Sample Name: S98T002999 SAM Date: 11/22/1998 11:06:00
 Data File : C:\DX\DATA\98112201.D19
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 19 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 51 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	145	686	2	
2	1.01	fluoride	14.842	862	5026	2	2.36
3	1.49	chloride	6.657	325	1520	1	0.22
4	1.77	nitrite	90.828	2622	14665	1	-0.93
5	2.89	nitrate	1836.223	38689	325625	1	-1.14
6	4.19	phosphate	44.042	225	2889	1	-0.16
7	5.49	sulfate	36.798	499	6581	1	0.06
8	7.15	oxalate	34.172	269	4823	1	-0.05

Totals 2063.562 43635 361815

File: 98112201.D19 Sample: S98T002995 SAM

1285

Sample Name: S98T00299 DUP Date: 11/22/1998 11:17:39
 Data File : C:\DX\DATA\98112201.D20
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 20 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	138	569	2	
2	1.01	fluoride	34.975	2114	11936	2	1.68
3	1.49	chloride	6.350	319	1445	1	-0.22
4	1.76	nitrite	83.336	2491	13406	1	-1.68
5	2.93	nitrate	672.707	15283	111363	1	0.23
6	4.19	phosphate	15.683	77	894	1	-0.32
7	5.49	sulfate	16.567	208	2702	1	0.06
8	7.15	oxalate	60.140	502	8564	1	-0.05
Totals			889.757	21131	150879		

File: 98112201.D20 Sample: S98T002995 DUP

Sample Name: S98T002995 SPK Date: 11/22/1998 13:38:19
 Data File : C:\DX\DATA\98112201.D21
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 21 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.06		0.000	109	188	1	
2	0.17		0.000	42	119	1	
3	0.69		0.000	32	62	1	
4	0.88		0.000	262	1195	2	
5	1.03	fluoride	45.216	2876	15500	2	3.70
6	1.51	chloride	49.428	2506	12164	1	1.12
7	1.80	nitrite	386.088	12052	64759	1	0.56
8	2.67	bromide	282.781	6185	33899	1	5.25
9	2.92	nitrate	2289.314	47366	418180	1	0.00
10	4.25	phosphate	325.862	1808	23151	1	1.27
11	5.55	sulfate	374.609	4917	72600	1	1.03
12	7.25	oxalate	318.685	2731	46202	1	1.45
Totals			4071.983	80886	688020		

File: 98112201.D21 Sample: S98T002995 SPK

Sample Name: S98T002996 SAM Date: 11/22/1998 13:50:05
 Data File : C:\DX\DATA\98112201.D22
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 22 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration $\mu\text{g/ml}$	Height	Area	Bl. Code	%Delta
1	0.87		0.000	144	662	2	
2	1.03	fluoride	11.304	693	3824	2	3.70
3	1.51	chloride	3.330	151	707	1	1.57
4	1.79	nitrite	39.659	1124	6077	1	0.19
5	2.91	nitrate	2249.656	46764	409829	1	-0.23
6	4.26	phosphate	43.883	214	2878	1	1.43
7	5.60	sulfate	26.351	376	4577	1	2.00
8	7.25	oxalate	36.618	285	5175	1	1.45
Totals			2410.801	49752	433729		

File: 98112201.D22 Sample: S98T002996 SAM

1238

Sample Name: S98T002996 DUP Date: 11/22/1998 14:02:42
 Data File : C:\DX\DATA\98112201.D23
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 23 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration $\mu\text{g/ml}$	Height	Area	Bl. Code	%Delta
1	0.87		0.000	128	544	2	
2	1.02	fluoride	15.120	925	5120	2	3.03
3	1.50	chloride	3.387	162	720	1	0.67
4	1.78	nitrite	38.891	1157	5948	1	-0.56
5	2.89	nitrate	2009.362	42557	360279	1	-1.14
6	4.23	phosphate	51.992	277	3450	1	0.63
7	5.55	sulfate	30.811	433	5432	1	1.03
8	7.20	oxalate	28.784	236	4048	1	0.70
Totals			2178.347	45875	385541		

File: 98112201.D23 Sample: S98T002996 DUP

1289

Sample Name: S98T002997 SAM Date: 11/22/1998 14:13:36
 Data File : C:\DX\DATA\98112201.D24
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 24 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 51 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	144	640	2	
2	1.02	fluoride	8.234	477	2784	2	3.03
3	1.49	chloride	6.110	284	1387	1	0.22
4	1.77	nitrite	67.012	1986	10664	1	-0.93
5	2.87	nitrate	2558.836	53764	476361	1	-1.83
6	4.22	phosphate	35.100	182	2259	1	0.48
7	5.55	sulfate	25.598	354	4433	1	1.03
8	7.20	oxalate	13.306	103	1822	1	0.70

Totals 2714.196 57294 500350

File: 98112201.D24 Sample: S98T002997 SAM

1290

Sample Name: S98T002997 DUP Date: 11/22/1998 14:28:24
 Data File : C:\DX\DATA\98112201.D25
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 25 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	151	718	2	
2	1.01	fluoride	8.091	467	2736	2	2.36
3	1.49	chloride	4.982	233	1110	1	0.22
4	1.77	nitrite	61.742	1799	9780	1	-0.93
5	2.85	nitrate	2656.910	55787	498189	1	-2.28
6	4.21	phosphate	29.333	148	1853	1	0.32
7	5.49	sulfate	30.190	353	5313	1	0.06
8	7.20	oxalate	12.640	110	1727	1	0.70
Totals			2803.888	59047	521425		

File: 98112201.D25 Sample: S98T002997 DUP

1291

LABCORE Completed Worklist Report for Worklist# 27277

Analyst: adp

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 (2), @IC4G-01 skm

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4 QC	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2 QC	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2 QC	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1 QC	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4 QC	QC	5.89e1	6.35e+01	107.810 %	Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2 QC	QC	1.24e2	1.13e+02	91.129 %	Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2 QC	QC	1.19e2	1.14e+02	95.798 %	Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLT1 QC	QC	1.07e2	1.05e+02	98.131 %	Recovery
3 CCV	0	@IC4G-QC F*4 QC	QC	6.21e1	5.82e+01	93.720 %	Recovery
3 CCV	0	@IC4G-QC ACETATE2 QC	QC	1.23e2	1.10e+02	89.431 %	Recovery
3 CCV	0	@IC4G-QC FORMATE2 QC	QC	1.30e2	1.19e+02	90.769 %	Recovery
3 CCV	0	@IC4G-QC GLYCOLT1 QC	QC	1.02e2	9.61e+01	94.216 %	Recovery
4 BLNK-PREP	0	@IC4G-01 F*4-01 SOLID	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2 SOLID	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2 SOLID	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLT1 SOLID	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002994	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	N/A	< 6.561e+01	65.640	ug/g
6 DUP	S98T002994	0 W @IC4G-01 F*4-01 SOLID	SOLID	1.65e+03	2.09e+03	23.529	RPD
6 DUP	S98T002994	0 W @IC4G-01 ACETATE2 SOLID	SOLID	2.47e+02	3.46e+02	33.390	RPD
6 DUP	S98T002994	0 W @IC4G-01 FORMATE2 SOLID	SOLID	1.00e+02	2.42e+02	83.041	RPD
6 DUP	S98T002994	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	<6.56e1	<6.60e1		RPD
7 SPK	S98T002994	0 W @IC4G-01 F*4-01 SOLID	SOLID	5.89e1	7.56e+01	128.353 %	Recovery
7 SPK	S98T002994	0 W @IC4G-01 ACETATE2 SOLID	SOLID	1.24e2	1.26e+02	101.613 %	Recovery
7 SPK	S98T002994	0 W @IC4G-01 FORMATE2 SOLID	SOLID	1.19e2	1.30e+02	109.244 %	Recovery
7 SPK	S98T002994	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	1.07e2	1.05e+02	98.131 %	Recovery
8 SAMPLE	S98T002995	0 W @IC4G-01 FORMATE2 SOLID	SOLID	N/A	2.932e+02	73.320	ug/g
8 SAMPLE	S98T002995	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	N/A	< 6.625e+01	66.240	ug/g
9 DUP	S98T002995	0 W @IC4G-01 F*4-01 SOLID	SOLID	2.89e+03	6.49e+03	76.759	RPD
9 DUP	S98T002995	0 W @IC4G-01 ACETATE2 SOLID	SOLID	4.56e+02	5.64e+02	21.176	RPD
9 DUP	S98T002995	0 W @IC4G-01 FORMATE2 SOLID	SOLID	2.93e+02	3.20e+02	8.809	RPD
9 DUP	S98T002995	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	<6.62e1	<6.56e1		RPD
10 SPK	S98T002995	0 W @IC4G-01 F*4-01 SOLID	SOLID	5.89e1	4.28e+01	72.666 %	Recovery
10 SPK	S98T002995	0 W @IC4G-01 ACETATE2 SOLID	SOLID	1.24e2	7.93e+01	60.726 %	Recovery
10 SPK	S98T002995	0 W @IC4G-01 FORMATE2 SOLID	SOLID	1.19e2	9.54e+01	80.168 %	Recovery
10 SPK	S98T002995	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	1.07e2	7.98e+01	74.579 %	Recovery
11 SAMPLE	S98T002996	0 W @IC4G-01 ACETATE2 SOLID	SOLID	N/A	2.291e+02	23.760	ug/g
11 SAMPLE	S98T002996	0 W @IC4G-01 FORMATE2 SOLID	SOLID	N/A	1.457e+02	73.620	ug/g
11 SAMPLE	S98T002996	0 W @IC4G-01 GLYCOLT1 SOLID	SOLID	N/A	< 6.652e+01	66.540	ug/g
12 DUP	S98T002996	0 W @IC4G-01 F*4-01 SOLID	SOLID	2.49e+03	3.15e+03	23.404	RPD
12 DUP	S98T002996	0 W @IC4G-01 ACETATE2 SOLID	SOLID	2.29e+02	2.09e+02	9.132	RPD
12 DUP	S98T002996	0 W @IC4G-01 FORMATE2 SOLID	SOLID	1.46e+02	1.53e+02	4.682	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27277

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
12 DUP	S98T002996	0 W	@IC4G-01 GLYCOLTI	SOLID	<6.65e1	<6.59e1		RPD
13 SAMPLE	S98T002997	0 W	@IC4G-01 F*4-01	SOLID	N/A	1.697e+03	36.610	ug/g
13 SAMPLE	S98T002997	0 W	@IC4G-01 ACETATE2	SOLID	N/A	3.236e+02	23.620	ug/g
13 SAMPLE	S98T002997	0 W	@IC4G-01 FORMATE2	SOLID	N/A	2.126e+02	73.200	ug/g
13 SAMPLE	S98T002997	0 W	@IC4G-01 GLYCOLTI	SOLID	N/A	6.614e+01	66.120	ug/g
14 DUP	S98T002997	0 W	@IC4G-01 F*4-01	SOLID	1.70e+03	1.65e+03		2.985 RPD
14 DUP	S98T002997	0 W	@IC4G-01 ACETATE2	SOLID	3.24e+02	3.22e+02		0.619 RPD
14 DUP	S98T002997	0 W	@IC4G-01 FORMATE2	SOLID	2.13e+02	1.99e+02		6.796 RPD
14 DUP	S98T002997	0 W	@IC4G-01 GLYCOLTI	SOLID	<6.61e1	<6.64e1		RPD

Final page for worklist# 27277

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

1/11/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27277

Analyst: ADD Instrument: IC Book# 3SN21-C LCS
Method: LA-533-1105 Rev/Mod F-0 34N21-A CCV
Worklist Comment: BY105 (2), @IC4G-01 skm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002994	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
6 DUP	S98T002994	0 W	@IC4G-01	SOLID		
7 SPK	S98T002994	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T002995	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
9 DUP	S98T002995	0 W	@IC4G-01	SOLID		
10 SPK	S98T002995	0 W	@IC4G-01	SOLID		
11 SAMPLE	S98T002996	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
12 DUP	S98T002996	0 W	@IC4G-01	SOLID		
13 SAMPLE	S98T002997	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
14 DUP	S98T002997	0 W	@IC4G-01	SOLID		

Final page for worklist # 27277

Anthony Perenta 12-31-98

Reanalysis for samples listed
below on WL# 27859. Blair
3/24/99

Data Entry Comments:

Faxed 12-31-98

some analytes on
S98T002994 & 2995

uploaded 1-4-99

John Howell 27277DEC.CSV

and For 2996 will
no run for high RPD.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1294

gm

Data Reprocessed On 12/31/1998 09:41:13

Sample Name: 35N21-C LCS Date: 12/31/1998 09:37:12
 Data File : C:\DX\DATA\98123101.D02
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2
 Analyst : *Anthony Peruto* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	8305	279284	2	
3	2.21		0.000	2326	17664	1	
7	2.77		0.000	1896	19465	2	
10	3.41		0.000	1620	22538	2	
12	3.84		0.000	2570	16663	2	
13	3.99		0.000	2975	36346	2	
14	4.20		0.000	2870	19680	2	
18	4.87		0.000	34864	1140312	2	
19	5.79		0.000	10138	264909	2	
20	6.44		0.000	4479	52429	2	
21	7.17	fluoride	63.546107	1920605	27304058	2	0.00
22	7.87	glycolate	105.3579846	530324	9381460	2	0.00
23	8.35	acetate	113.1629126	593060	11330272	2	0.00
24	9.04	formate	113.919957	8029578	19129228	2	0.00
25	11.33		0.000	8793154	407940685	2	
26	12.05		0.000	8015525	640992672	2	

Totals 395.983 20954288 1117947666

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 completed / VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1296 TO 1314.

1295

Data Reprocessed On 12/31/1998 10:00:13

Sample Name: 34N21-A CCV	Date: 12/31/1998 09:57:18
Data File : C:\DX\DATA\98123101.D03	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 3	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	14028	782166	1	
7	2.87		0.000	3213	23211	2	
8	3.04		0.000	2615	16707	2	
9	3.17		0.000	2423	34433	2	
10	3.55		0.000	3196	32212	2	
11	3.71		0.000	1993	17893	2	
14	4.45		0.000	2369	20107	2	
25	5.91		0.000	1275	31728	1	
27	7.17	fluoride	58.191	1707895	24937633	2	0.00
28	7.89	glycolate	96.074	482227	8551860	2	0.00
29	8.37	acetate	110.484	578848	11068298	2	0.00
30	9.07	formate	118.077	1058823	19840184	2	0.00
32	10.85		0.000	1639678	25122577	2	
33	11.33		0.000	8690565	361326796	2	
34	12.05		0.000	7490894	638512958	3	
35	15.84		0.000	30089	1126961	4	

Totals	382.825	21710131	1091445726

1236

```

=====
Sample Name: BLANK                      Date: 12/31/1998 09:15:40
Data File  : C:\DX\DATA\98123101.D01
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 1      Detector: CDM-2
Analyst    :                          Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 5700 5Hz 0.00 19.00 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	39631	502218	1	
2	2.03		0.000	58010	870800	2	
3	2.40		0.000	25400	299897	2	
5	2.99		0.000	3873	22200	1	
6	3.53		0.000	4871	115898	2	
8	4.16		0.000	5190	113592	3	
10	4.36		0.000	49459	865512	2	
11	4.55		0.000	175823	3300648	2	
12	5.72		0.000	1948853	103990817	2	
13	6.93		0.000	3202819	238335874	2	
14	9.09		0.000	3037	17600	1	
16	11.33		0.000	8753443	397650935	2	
17	12.05		0.000	11714013	1274037426	3	
18	18.67		0.000	3468	21850	4	
Totals			0.000	25987887	2020145268		

File: 98123101.D01 Sample: BLANK

1297

Data Reprocessed On 01/04/1999 13:51:46

```

=====
Sample Name: PREP BLANK                      Date: 12/31/1998 10:33:54
Data File  : E:\DATA\98123101.D04
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 4          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1  5700  5Hz  0.00 19.00 25000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.50		0.000	41621	2834400	1	
6	11.33		0.000	8604639	438721150	2	
7	12.05		0.000	8291175	578384416	2	
Totals			0.000	16937435	1019939966		

File: 98123101.D04 Sample: PREP BLANK

Data Reprocessed On 01/04/1999 13:57:09

```

=====
Sample Name: S98T002994 SAM                      Date: 12/31/1998 11:03:49
Data File  : E:\DATA\98123101.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	5700	5Hz	6.38	11.59	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	6618	250926	2	
2	1.70		0.000	109935	1298291	2	
3	2.89		0.000	10756	179799	2	
4	2.99		0.000	10636	85037	2	
5	3.13		0.000	9693	63211	2	
6	3.24		0.000	9267	73985	2	
7	3.41		0.000	9072	133646	2	
8	3.68		0.000	7978	52360	2	
9	3.81		0.000	7241	41991	2	
10	3.99		0.000	7127	94027	2	
11	4.16		0.000	5712	44257	2	
12	4.39		0.000	4755	44196	2	
13	4.48		0.000	4755	44991	2	
14	4.68		0.000	5349	77046	2	
15	4.92		0.000	3666	27147	2	
16	5.13		0.000	3015	27087	2	
17	5.35		0.000	2923	33033	2	
19	5.56		0.000	2952	21577	2	
22	7.07	fluoride	8.461	4262005	62996503	3	0.00
24	8.21	acetate	1.266	136625	2320895	2	0.00
25	8.93	formate	0.514	94568	1390000	1	0.00
Totals			10.241	4714645	69300004		

Data Reprocessed On 01/04/1999 14:02:58

```

=====
Sample Name: S98T002994 DUP                      Date: 12/31/1998 11:23:57
Data File  : E:\DATA\98123101.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 6.39 10.86 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	4564	180556	1	
2	1.70		0.000	111243	1278174	1	
3	2.97		0.000	12595	295384	2	
4	3.12		0.000	12656	101774	2	
5	3.27		0.000	12657	144357	2	
6	3.63		0.000	11871	199268	2	
7	3.77		0.000	11852	95205	2	
8	3.91		0.000	12455	124951	2	
9	4.17		0.000	11800	185741	2	
10	4.43		0.000	11548	83507	2	
11	4.52		0.000	11276	81644	2	
12	4.63		0.000	11282	125354	2	
13	4.84		0.000	10915	52339	2	
14	4.93		0.000	11322	54825	2	
15	5.01		0.000	11812	99808	2	
16	5.17		0.000	10312	180382	2	
17	5.52		0.000	8927	79240	2	
18	5.68		0.000	9226	44342	2	
19	5.75		0.000	9338	177432	2	
20	7.09	fluoride	10.644	5200794	80202985	3	0.00
21	7.84	glycolate	0.014	18341	188670	4	0.00
22	8.27	acetate	1.759	148663	3137942	2	0.00
23	8.96	formate	1.234	172854	3407528	2	0.00
24	9.73		0.000	35725	719661	2	
25	10.88		0.000	5301064	53262494	2	
26	11.15		0.000	34730446	389666966	2	
27	11.65		0.000	428868972	7466345865	2	
Totals			13.651	474784509	8000516397		

```

=====
Sample Name: S98T002994 SPK                      Date: 12/31/1998 11:45:04
Data File  : E:\DATA\98123101.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 6.19 10.41 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.37		0.000	7928	148205	2	
4	1.70		0.000	107699	1281768	1	
5	3.15		0.000	4616	248664	3	
13	4.83		0.000	1638	25462	2	
14	5.15		0.000	2768	58873	2	
15	5.51		0.000	3797	37115	2	
16	5.68		0.000	4698	90550	2	
18	7.09	fluoride	12.241	5954146	93037086	2	0.00
19	7.79	glycolate	5.253	425971	7870975	2	0.00
20	8.29	acetate	7.586	660777	12750904	2	0.00
21	8.96	formate	6.998	1045327	19793135	3	0.00
22	9.73		0.000	37054	692631	4	
23	10.88		0.000	5584737	54708284	2	
24	11.15		0.000	35018979	419758168	2	
25	11.65		0.000	392251540	6684184732	2	
Totals			32.078	441111675	7294686551		

File: 98123101.D07 Sample: S98T002994 SPK

```

=====
Sample Name: S98T002995 SAM                      Date: 12/31/1998 12:05:42
Data File  : E:\DATA\98123101.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   6.13 10.90    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.37		0.000	12826	276448	1	
3	1.70		0.000	112684	1378066	1	
4	2.77		0.000	6984	90569	2	
5	2.89		0.000	5583	51176	2	
8	3.19		0.000	2045	21502	1	
15	4.44		0.000	2956	19285	2	
17	4.81		0.000	1771	35485	2	
18	5.12		0.000	2962	59337	2	
19	5.52		0.000	5271	130113	2	
20	5.96		0.000	4911	44219	2	
21	7.09	fluoride	14.669	7236680	112987237	3	0.00
22	7.79	glycolate	0.066	25040	263884	4	0.00
23	8.27	acetate	2.313	194997	4054494	2	0.00
24	8.96	formate	1.487	211899	4116099	2	0.00
25	9.71		0.000	40736	832718	2	
26	10.88		0.000	6334819	57853862	2	
27	11.15		0.000	41508455	462349916	2	
28	11.68		0.000	321691728	5260746033	3	
29	12.29		0.000	6116044	374146218	4	
30	17.84		0.000	6000	68400	4	
31	18.45		0.000	1143	16800	1	
Totals			18.535	383525533	6279541862		

```

=====
Sample Name: S98T002995 DUP                      Date: 12/31/1998 12:25:21
Data File  : E:\DATA\98123101.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                               Detector: CDM-2
                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   5.72 10.65    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	7568	35686	2	
2	0.37		0.000	7696	95394	2	
4	1.70		0.000	124278	1442106	1	
5	2.84		0.000	17594	261997	2	
6	2.97		0.000	23828	279003	2	
7	3.17		0.000	18840	155775	2	
8	3.35		0.000	17365	66723	2	
9	3.41		0.000	18603	103046	2	
10	3.52		0.000	16691	176695	2	
11	3.75		0.000	16530	214264	2	
12	3.96		0.000	15465	75937	2	
13	4.05		0.000	14689	93128	2	
14	4.24		0.000	13738	126972	2	
15	4.36		0.000	13369	83764	2	
16	4.53		0.000	15354	107437	2	
17	4.60		0.000	16292	123956	2	
18	4.75		0.000	13270	76234	2	
19	4.85		0.000	13237	84231	2	
20	4.99		0.000	12972	92241	2	
21	5.12		0.000	12747	111134	2	
22	5.28		0.000	12468	202272	2	
23	5.61		0.000	11255	277106	2	
24	7.09	fluoride	32.780	17070335	283991784	2	0.00
25	8.27	acetate	2.851	236157	4943987	2	0.00
26	8.96	formate	1.619	231292	4489126	2	0.00
27	9.71		0.000	38065	772968	2	
28	10.88		0.000	6236803	56903865	2	
29	11.15		0.000	43667408	478017447	2	
30	11.79		0.000	160840461	2337203833	3	
31	12.24		0.000	12138582	462874811	4	
32	18.67		0.000	2000	801580	1	
Totals			37.251	240894952	3634284503		

Data Reprocessed On 01/04/1999 14:22:04

Sample Name: S98T002995 SPK	Date: 12/31/1998 12:59:50
Data File : E:\DATA\98123101.D10	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 10	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	6	5700	5Hz	6.23	11.39	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	2803	76125	3	
3	1.70		0.000	127929	1699207	2	
4	7.04	fluoride	16.809	8647885	131039673	2	0.00
5	7.71	glycolate	3.989	350506	5990302	2	0.00
6	8.21	acetate	6.077	554177	10267435	2	0.00
7	8.88	formate	6.257	978298	17662996	2	0.00
8	9.60		0.000	36768	677221	2	
9	10.85		0.000	6869509	92312521	2	
10	11.15		0.000	38545556	469474313	2	
11	11.71		0.000	341007010	5179176212	3	
12	12.29		0.000	6515084	387668792	4	

Totals	33.133	403635526	6296044797

File: 98123101.D10 Sample: S98T002995 SPK

Data Reprocessed On 01/04/1999 14:23:39

```

=====
Sample Name: S98T002996 SAM                      Date: 12/31/1998 13:19:52
Data File  : E:\DATA\98123101.D11
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 11
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 6.40 10.74 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.33		0.000	4903	87345	2	
5	1.70		0.000	113188	1316108	1	
6	2.73		0.000	6866	82989	2	
7	2.88		0.000	4684	42633	2	
13	4.05		0.000	2210	19857	2	
15	4.47		0.000	2173	18020	2	
26	7.09	fluoride	12.591	6186892	95880724	2	0.00
27	8.27	acetate	1.157	96788	2140156	2	0.00
28	8.96	formate	0.736	101059	2009599	2	0.00
29	9.73		0.000	19914	399491	2	
30	10.88		0.000	3534158	44034734	2	
31	11.15		0.000	21495505	321345471	2	
32	11.65		0.000	380273963	6222237096	3	
33	12.29		0.000	7339550	393072789	4	
Totals			14.483	419181853	7082687011		

File: 98123101.D11 Sample: S98T002996 SAM

```

=====
Sample Name: S98T002996 DUP                      Date: 12/31/1998 13:43:06
Data File  : E:\DATA\98123101.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          6    5700  5Hz   6.24 10.76    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.43		0.000	8402	154928	2	
3	1.70		0.000	111473	1284143	2	
4	2.75		0.000	11775	182566	2	
5	2.91		0.000	12456	150675	2	
6	3.11		0.000	13781	187928	2	
7	3.43		0.000	12741	156241	2	
8	3.69		0.000	11648	113028	2	
9	3.76		0.000	12130	92852	2	
10	3.93		0.000	11459	156565	2	
11	4.16		0.000	9901	68800	2	
12	4.35		0.000	10137	92309	2	
13	4.47		0.000	9193	82697	2	
14	4.69		0.000	9075	68755	2	
15	4.81		0.000	10330	121698	2	
16	5.12		0.000	8322	107546	2	
17	5.27		0.000	8294	84901	2	
18	5.51		0.000	7904	54203	2	
19	5.63		0.000	8339	141765	3	
21	7.09	fluoride	16.057	8040180	124641127	2	0.00
22	8.27	acetate	1.068	93605	1993182	2	0.00
23	8.93	formate	0.778	107462	2127858	2	0.00
24	9.68		0.000	20804	422580	2	
25	10.88		0.000	3826248	36781418	2	
26	11.15		0.000	21212155	357549322	2	
27	11.68		0.000	361467967	5666092114	3	
28	12.29		0.000	6103186	381801785	4	
Totals			17.903	401158967	6574710985		

Data Reprocessed On 01/04/1999 14:27:01

```

=====
Sample Name: S98T002997 SAM                      Date: 12/31/1998 14:03:23
Data File  : E:\DATA\98123101.D13
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 13              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          6    5700  5Hz   6.28 10.72   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	2983	25997	2	
4	1.70		0.000	115066	1414826	1	
5	2.83		0.000	7342	130588	2	
6	2.99		0.000	5387	48020	2	
7	3.13		0.000	3375	32954	2	
11	3.85		0.000	2521	30440	2	
16	4.63		0.000	2519	24058	2	
17	4.92		0.000	2624	28105	2	
22	5.61		0.000	2504	25552	2	
23	5.85		0.000	3381	38154	2	
24	7.09	fluoride	8.623	4211622	64263109	3	0.00
25	7.79	glycolate	0.084	17988	290068	4	0.00
26	8.27	acetate	1.644	146120	2946232	2	0.00
27	8.96	formate	1.080	151927	2973536	2	0.00
28	9.73		0.000	33016	677579	2	
29	10.88		0.000	5112542	51468965	2	
30	11.15		0.000	32849236	385887556	2	
31	11.65		0.000	431866453	7494034312	2	
32	18.53		0.000	1413	20840	1	
Totals			11.430	474538019	8004360891		

Data Reprocessed On 01/04/1999 14:33:08

Sample Name: S98T002997 DUP	Date: 12/31/1998 14:23:02
Data File : E:\DATA\98123101.D14	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 14	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	5700	5Hz	6.39	10.29	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.13		0.000	3507	17941	2	
2	0.40		0.000	3906	70440	2	
3	1.70		0.000	117719	1411532	1	
4	2.75		0.000	9477	97470	2	
5	2.92		0.000	10254	249766	2	
6	3.31		0.000	6724	78357	2	
7	3.56		0.000	3336	31334	2	
8	3.76		0.000	2364	22090	2	
17	5.57		0.000	3747	120207	2	
18	7.09	fluoride	8.362	4085835	62225379	3	0.00
19	7.76	glycolate	0.031	17776	214068	4	0.00
20	8.27	acetate	1.630	139992	2924564	2	0.00
21	8.96	formate	1.009	142072	2776752	2	0.00
22	9.73		0.000	30164	630833	2	
23	10.88		0.000	4704325	53650446	2	
24	11.15		0.000	29986914	357884517	2	
25	11.63		0.000	437887776	7869581457	2	
Totals			11.033	477155888	8351987153		

File: 98123101.D14 Sample: S98T002997 DUP

1314 Minutes

LABCORE Completed Worklist Report for Worklist# 27291

Analyst: adp

Instrument: IC40S1

Book#: 29N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105(2) @IC-01 rerun. Use 101 df for NO3 and 41df for rest

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1	CCB	0	QIC-QC	F	QC	1	<1.20e-2	ug/mL	
1	CCB	0	QIC-QC	CL	QC	1	<1.70e-2	ug/mL	
1	CCB	0	QIC-QC	NO2	QC	1	<1.08e-1	ug/mL	
1	CCB	0	QIC-QC	BR	QC	1	<1.25e-1	ug/mL	
1	CCB	0	QIC-QC	NO3	QC	1	<1.39e-1	ug/mL	
1	CCB	0	QIC-QC	PO4	QC	1	<1.20e-1	ug/mL	
1	CCB	0	QIC-QC	SO4	QC	1	<1.39e-1	ug/mL	
1	CCB	0	QIC-QC	OXALATE2	QC	1	<1.05e-1	ug/mL	
2	LCS-INST	0	QIC-QC	F	QC	5.90e1	6.67e+01	113.051 % Recovery	
2	LCS-INST	0	QIC-QC	CL	QC	8.00e1	8.43e+01	105.375 % Recovery	
2	LCS-INST	0	QIC-QC	NO2	QC	5.48e2	5.66e+02	103.285 % Recovery	
2	LCS-INST	0	QIC-QC	BR	QC	5.86e2	5.96e+02	101.706 % Recovery	
2	LCS-INST	0	QIC-QC	NO3	QC	5.92e2	6.08e+02	102.703 % Recovery	
2	LCS-INST	0	QIC-QC	PO4	QC	5.47e2	5.49e+02	100.366 % Recovery	
2	LCS-INST	0	QIC-QC	SO4	QC	6.38e2	6.64e+02	104.075 % Recovery	
2	LCS-INST	0	QIC-QC	OXALATE2	QC	5.40e2	5.59e+02	103.519 % Recovery	
3	CCV	0	QIC-QC	F	QC	6.40e1	6.87e+01	107.344 % Recovery	
3	CCV	0	QIC-QC	CL	QC	9.00e1	9.40e+01	104.444 % Recovery	
3	CCV	0	QIC-QC	NO2	QC	5.62e2	5.74e+02	102.135 % Recovery	
3	CCV	0	QIC-QC	BR	QC	6.30e2	6.28e+02	99.683 % Recovery	
3	CCV	0	QIC-QC	NO3	QC	6.98e2	7.27e+02	104.155 % Recovery	
3	CCV	0	QIC-QC	PO4	QC	6.32e2	6.66e+02	105.380 % Recovery	
3	CCV	0	QIC-QC	SO4	QC	6.99e2	7.24e+02	103.577 % Recovery	
3	CCV	0	QIC-QC	OXALATE2	QC	5.53e2	5.83e+02	105.425 % Recovery	
4	BLNK-PREP	0	QIC-01	F-02	SOLID	1	<1.20e-2	ug/g	
4	BLNK-PREP	0	QIC-01	CL-02	SOLID	1	<1.70e-2	ug/g	
4	BLNK-PREP	0	QIC-01	NO2-02	SOLID	1	6.38e-01	0.638 ug/g	
4	BLNK-PREP	0	QIC-01	BR-02	SOLID	1	<1.25e-1	ug/g	
4	BLNK-PREP	0	QIC-01	NO3-02	SOLID	1	1.79e-01	0.179 ug/g	
4	BLNK-PREP	0	QIC-01	PO4-02	SOLID	1	<1.20e-1	ug/g	
4	BLNK-PREP	0	QIC-01	SO4-02	SOLID	1	<1.38e-1	ug/g	
4	BLNK-PREP	0	QIC-01	OXALATE2	SOLID	1	<1.05e-1	ug/g	
5	SAMPLE	S98T002994	0 W	QIC-01	NO3-02	SOLID	N/A	4.240e+05	2741.000 ug/g
6	DUP	S98T002994	0 W	QIC-01	F-02	SOLID	?	2.48e+03	RPD
6	DUP	S98T002994	0 W	QIC-01	CL-02	SOLID	?	1.04e+03	RPD
6	DUP	S98T002994	0 W	QIC-01	NO2-02	SOLID	?	1.34e+04	RPD
6	DUP	S98T002994	0 W	QIC-01	BR-02	SOLID	?	<1.01e3	RPD
6	DUP	S98T002994	0 W	QIC-01	NO3-02	SOLID	4.24e+05	4.78e+05	11.973 RPD
6	DUP	S98T002994	0 W	QIC-01	PO4-02	SOLID	?	6.53e+03	RPD
6	DUP	S98T002994	0 W	QIC-01	SO4-02	SOLID	?	5.54e+03	RPD
6	DUP	S98T002994	0 W	QIC-01	OXALATE2	SOLID	?	2.31e+03	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27291

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 SPK	S98T002994	0 W	@IC-01	F-02	SOLID	5.90e1	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	CL-02	SOLID	8.00e1	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	NO2-02	SOLID	5.48e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	BR-02	SOLID	5.86e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	NO3-02	SOLID	5.92e2	7.03e+02	118.750 % Recovery
7 SPK	S98T002994	0 W	@IC-01	PO4-02	SOLID	5.47e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	SO4-02	SOLID	5.38e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002994	0 W	@IC-01	OXALATE2	SOLID	5.40e2	0.00e+00	0.000 % Recovery
8 SAMPLE	S98T002996	0 W	@IC-01	F-02	SOLID	N/A	2.748e+03	97.420 ug/g
8 SAMPLE	S98T002996	0 W	@IC-01	OXALATE2	SOLID	N/A	6.464e+03	852.400 ug/g
9 DUP	S98T002996	0 W	@IC-01	F-02	SOLID	2.75e+03	3.35e+03	19.672 RPD
9 DUP	S98T002996	0 W	@IC-01	CL-02	SOLID	?	5.34e+02	RPD
9 DUP	S98T002996	0 W	@IC-01	NO2-02	SOLID	?	7.46e+03	RPD
9 DUP	S98T002996	0 W	@IC-01	BR-02	SOLID	?	<1.01e3	RPD
9 DUP	S98T002996	0 W	@IC-01	NO3-02	SOLID	?	3.80e+05	RPD
9 DUP	S98T002996	0 W	@IC-01	PO4-02	SOLID	?	1.00e+04	RPD
9 DUP	S98T002996	0 W	@IC-01	SO4-02	SOLID	?	5.73e+03	RPD
9 DUP	S98T002996	0 W	@IC-01	OXALATE2	SOLID	6.46e+03	6.03e+03	6.886 RPD

Final page for worklist# 27291

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11/30/98

LABCORE Data Entry Template for Worklist# 27291

Analyst: AMP Instrument: IC 4051 Book# 29N21-B LCS
Method: LA-533-105 Rev/Mod F-0 28N21-B CCV

Worklist Comment: BY-105(2) @IC-01 rerun. Use 101 df for NO3 and 41df for rest

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T002994 0 W			@IC-01	SOLID	98000509	BY-105 (2)
		Analytes Requested: NO3-02						
6	DUP	S98T002994 0 W			@IC-01	SOLID		
7	SPK	S98T002994 0 W			@IC-01	SOLID		
8	SAMPLE	S98T002996 0 W			@IC-01	SOLID	98000509	BY-105 (2)
		Analytes Requested: F-02					, OXALATE2	
9	DUP	S98T002996 0 W			@IC-01	SOLID		

Final page for worklist # 27291

11-25-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 11-30-98
John Wawell
27291/NOV.CSV

Validated 11/30/98 mtf

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1317

Sample Name: 29N21-B LCS Date: 11/25/1998 19:18:52
 Data File : C:\DX\DATA\98112521.D02
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: System: 1 Inject#: 2 Detector: CDM-1
 Analyst : *Anthony Purnell* Column: AG4A/AS4A anion column

11-25-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.88		0.000	67	272	2	
2	1.03	fluoride	66.733	113.11	2083	2	3.70
3	1.49	chloride	84.332	105.42	2184	1	0.22
4	1.78	nitrite	566.170	103.32	9014	1	-0.56
5	2.61	bromide	595.712	101.66	6062	1	2.62
6	2.95	nitrate	607.980	102.70	7127	1	0.91
7	4.25	phosphate	549.412	100.44	1598	1	1.27
8	5.55	sulfate	664.450	105.49	5259	1	1.03
9	7.20	oxalate	558.975	103.14	2482	1	0.70
Totals			3693.764	35877	280784		

File: 98112521.D02 Sample: 29N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1318 TO 1331.

1318

Sample Name: 28N21-~~2~~6 CCV
 Data File : C:\DX\DATA\98112521.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3
 Analyst : Column: AG4A/AS4A anion column
 Date: 11/25/1998 19:51:55
 Detector: CDM-1

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	84	357	2	
2	1.01	fluoride	68.694	2168	11836	2	2.36
3	1.47	chloride	94.024	2468	11671	1	-1.12
4	1.75	nitrite	574.232	9182	48347	1	-2.05
5	2.58	bromide	627.504	6696	38112	1	1.57
6	2.91	nitrate	727.451	8754	59553	1	-0.46
7	4.21	phosphate	666.335	1905	23929	1	0.32
8	5.49	sulfate	723.652	5758	70769	1	0.06
9	7.15	oxalate	30583.141	2629	42644	1	-0.05
Totals			4065.032	39644	307219		

File: 98112521.D03 Sample: 28N21-~~2~~6 CCV

1319

Sample Name: BLANK Date: 11/25/1998 19:04:44
Data File : C:\DX\DATA\98112521.D01
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 1 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
Totals			0.000	0	0		

=====

Sample Name:	PREP BLANK	Date:	11/25/1998 20:13:39
Data File :	C:\DX\DATA\98112521.D04		
Method :	C:\DX\METHOD\KIT.MET		
ACI Address:	1	System:	1
Inject#:	4	Detector:	CDM-1
Analyst :		Column:	AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.31		0.000	58	542	1	
2	1.77	nitrite	0.638	814	4885	1	-1.30
3	2.97	nitrate	0.179	120	712	1	1.60
Totals			0.818	992	6138		

File: 98112521.D04 Sample: PREP BLANK

Sample Name: S98T002994 SAM Date: 11/25/1998 20:29:25
 Data File : C:\DX\DATA\98112521.D05
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 5 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	48	200	2	
2	1.01	fluoride	7.903	245	1352	2	2.36
3	1.48	chloride	5.786	136	607	1	-0.67
4	1.76	nitrite	72.017	1069	5525	1	-1.68
5	2.87	nitrate	2171.268	24863	186036	1	-1.60
6	4.23	phosphate	28.633	70	808	1	0.79
7	5.55	sulfate	29.894	190	2420	1	1.03
8	7.20	oxalate	10.858	46	698	1	0.70
Totals			2326.359	26666	197646		

File: 98112521.D05 Sample: S98T002994 SAM

```

=====
Sample Name: S98T002994 SAM                      Date: 11/25/1998 20:52:59
Data File  : C:\DX\DATA\98112521.D07
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 7              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	152	701	2	
2	1.01	fluoride	10.576	655	4452	2	2.36
3	1.47	chloride	5.476	316	1558	1	-1.12
4	1.75	nitrite	72.543	2652	14565	1	-2.05
5	2.81	nitrate	2193.576	58675	514382	1	-3.88
6	4.22	phosphate	27.803	186	2223	1	0.48
7	5.49	sulfate	25.158	392	5525	1	0.06
8	7.15	oxalate	10.671	101	1818	1	-0.05
Totals			2345.803	63129	545224		

File: 98112521.D07 Sample: S98T002994 SAM

1323

Sample Name: S98T002994 DUP Date: 11/25/1998 20:40:49
 Data File : C:\DX\DATA\98112521.D06
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 6 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.88		0.000	55	254	2	
2	1.02	fluoride	14.129	321	2412	2	3.03
3	1.49	chloride	4.877	117	495	1	-0.22
4	1.76	nitrite	67.752	990	5164	1	-1.68
5	2.87	nitrate	2434.704	27963	210216	1	-1.60
6	4.25	phosphate	37.257	92	1114	1	1.11
7	5.55	sulfate	32.983	213	2719	1	1.03
8	7.20	oxalate	12.459	50	815	1	0.70
Totals			2604.161	29801	223189		

File: 98112521.D06 Sample: S98T002994 DUP

1324

Data Reprocessed On 11/30/1998 13:33:16

```

=====
Sample Name: S98T002994 DUP                      Date: 11/25/1998 21:03:50
Data File  : F:\DATA\98112521.D08
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 8                Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	152	691	2	
2	1.01	fluoride	12.605	780	5310	2	2.36
3	1.47	chloride	5.309	308	1507	1	-1.12
4	1.75	nitrite	68.279	2509	13674	1	-2.05
5	2.80	nitrate	2407.995	64852	576552	1	0.00
6	4.23	phosphate	33.257	222	2701	1	0.63
7	5.49	sulfate	28.229	453	6258	1	0.06
8	7.15	oxalate	11.739	113	2009	1	-0.05
Totals			2567.413	69389	608702		

File: 98112521.D08 Sample: S98T002994 DUP

Sample Name: S98T002994 SPK Date: 11/25/1998 21:17:40
 Data File : C:\DX\DATA\98112521.D09
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 9 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	116	494	2	
2	1.01	fluoride	72.443	2256	12493	2	2.36
3	1.48	chloride	90.799	2279	11260	1	-0.67
4	1.76	nitrite	647.087	10096	54616	1	-1.68
5	2.58	bromide	570.296	6216	34539	1	1.57
6	2.85	nitrate	2874.603	32484	251448	1	-2.28
7	4.22	phosphate	575.715	1657	20579	1	0.48
8	5.49	sulfate	676.740	5407	66064	1	0.06
9	7.15	oxalate	570.499	2548	41708	1	-0.05
Totals			6078.182	63060	493201		

File: 98112521.D09 Sample: S98T002994 SPK

1326

Data Reprocessed On 11/30/1998 13:34:49

```

=====
Sample Name: S98T002994 SPK                      Date: 11/25/1998 22:09:29
Data File  : F:\DATA\98112521.D11
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 11              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      41    3000  5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	210	885	2	
2	1.01	fluoride	35.919	2698	15313	2	2.36
3	1.47	chloride	39.198	2387	11995	1	-1.12
4	1.75	nitrite	309.474	11581	64565	1	-2.05
5	2.57	bromide	222.285	6363	33124	1	1.05
6	2.79	nitrate	2394.738	64490	572615	1	0.12
7	4.21	phosphate	249.614	1773	22025	1	0.16
8	5.49	sulfate	288.055	5799	69358	1	0.06
9	7.09	oxalate	231.853	2400	41756	1	-0.79
Totals			3771.137	97700	831637		

File: 98112521.D11 Sample: S98T002994 SPK

```

=====
Sample Name: S98T002996 SAM                      Date: 11/25/1998 21:58:17
Data File  : C:\DX\DATA\98112521.D10
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 10              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz  0.00 10.00      30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	58	266	2	
2	1.02	fluoride	16.340	375	2790	2	3.03
3	1.49	chloride	3.695	81	349	1	-0.22
4	1.76	nitrite	40.204	547	2835	1	-1.68
5	2.87	nitrate	2143.341	24783	183494	1	-1.60
6	4.23	phosphate	43.411	103	1332	1	0.79
7	5.49	sulfate	27.968	157	2234	1	0.06
8	7.15	oxalate	36.160	137	2534	1	-0.05
Totals			2311.118	26239	195835		

File: 98112521.D10 Sample: S98T002996 SAM

1328

```

=====
Sample Name: S98T002996 SAM                      Date: 11/25/1998 23:00:46
Data File  : C:\DX\DATA\98112521.D15
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 15              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           41    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	137	623	2	
2	1.02	fluoride	13.882	928	5852	2	3.03
3	1.48	chloride	2.878	163	768	1	-0.67
4	1.76	nitrite	38.445	1354	7445	1	-1.68
5	2.81	nitrate	2163.244	57586	505833	1	-3.65
6	4.23	phosphate	38.817	263	3189	1	0.63
7	5.49	sulfate	25.338	394	5568	1	0.06
8	7.15	oxalate	32.648	336	5750	1	-0.05
Totals			2315.252	61161	535027		

File: 98112521.D15 Sample: S98T002996 SAM

1329

=====

Sample Name: S98T002996 DUP	Date: 11/25/1998 22:33:17
Data File : C:\DX\DATA\98112521.D13	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 13	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	53	233	2	
2	1.01	fluoride	20.107	493	3434	2	2.36
3	1.47	chloride	3.316	73	302	1	-1.12
4	1.75	nitrite	42.408	577	3021	1	-2.42
5	2.86	nitrate	1936.076	22434	164755	1	-2.05
6	4.21	phosphate	50.433	133	1581	1	0.32
7	5.49	sulfate	32.181	208	2642	1	0.06
8	7.15	oxalate	24.611	112	1696	1	-0.05
Totals			2109.132	24082	177665		

File: 98112521.D13 Sample: S98T002996 DUP

Sample Name: S98T002996 DUP Date: 11/25/1998 22:46:12
 Data File : C:\DX\DATA\98112521.D14
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 14 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	41	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	126	548	2	
2	1.02	fluoride	17.089	1202	7214	2	3.03
3	1.47	chloride	2.720	159	720	1	-1.12
4	1.75	nitrite	38.012	1377	7355	1	-2.05
5	2.81	nitrate	1925.168	51774	440632	1	-3.65
6	4.22	phosphate	51.234	344	4279	1	0.48
7	5.49	sulfate	29.233	484	6497	1	0.06
8	7.15	oxalate	30.740	296	5408	1	-0.05
9	8.91		0.000	16	550	1	
Totals			2094.196	55777	473204		

File: 98112521.D14 Sample: S98T002996 DUP

LABCORE Completed Worklist Report for Worklist# 27292

Analyst: vlm

Instrument: IC40S2

Book#: 29N21B

Method: LA-533-105 Rev/Mod F--O

Worklist Comment: BY-105(2) @ic-01 rerun. Use 101df for NO3 and 41df for rest.

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	0	01C-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	0	01C-QC CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	0	01C-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	0	01C-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	0	01C-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	0	01C-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	0	01C-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	0	01C-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	0	01C-QC F	QC	5.90e1	6.45e+01	109.322 % Recovery	
2 LCS-INST	0	0	01C-QC CL	QC	8.00e1	8.41e+01	105.125 % Recovery	
2 LCS-INST	0	0	01C-QC NO2	QC	5.48e2	5.33e+02	97.263 % Recovery	
2 LCS-INST	0	0	01C-QC BR	QC	5.86e2	6.16e+02	105.119 % Recovery	
2 LCS-INST	0	0	01C-QC NO3	QC	5.32e2	6.30e+02	106.419 % Recovery	
2 LCS-INST	0	0	01C-QC P04	QC	5.47e2	5.44e+02	99.452 % Recovery	
2 LCS-INST	0	0	01C-QC S04	QC	6.38e2	6.67e+02	104.545 % Recovery	
2 LCS-INST	0	0	01C-QC OXALATE2	QC	5.40e2	5.66e+02	104.815 % Recovery	
3 CCV	0	0	01C-QC F	QC	6.40e1	6.49e+01	101.406 % Recovery	
3 CCV	0	0	01C-QC CL	QC	9.00e1	9.53e+01	105.889 % Recovery	
3 CCV	0	0	01C-QC NO2	QC	5.62e2	5.50e+02	97.865 % Recovery	
3 CCV	0	0	01C-QC BR	QC	6.30e2	6.82e+02	108.254 % Recovery	
3 CCV	0	0	01C-QC NO3	QC	6.98e2	7.54e+02	108.023 % Recovery	
3 CCV	0	0	01C-QC P04	QC	6.32e2	6.33e+02	100.158 % Recovery	
3 CCV	0	0	01C-QC S04	QC	6.99e2	7.17e+02	102.575 % Recovery	
3 CCV	0	0	01C-QC OXALATE2	QC	5.53e2	5.84e+02	105.606 % Recovery	
4 BLNK-PREP	0	0	01C-01 F-02	SOLID	1	<1.20e-2		ug/g
4 BLNK-PREP	0	0	01C-01 CL-02	SOLID	1	3.40e-02	0.034	ug/g
4 BLNK-PREP	0	0	01C-01 NO2-02	SOLID	1	4.99e-01	0.499	ug/g
4 BLNK-PREP	0	0	01C-01 BR-02	SOLID	1	<1.25e-1		ug/g
4 BLNK-PREP	0	0	01C-01 NO3-02	SOLID	1	1.55e-01	0.155	ug/g
4 BLNK-PREP	0	0	01C-01 P04-02	SOLID	1	<1.20e-1		ug/g
4 BLNK-PREP	0	0	01C-01 S04-02	SOLID	1	<1.38e-1		ug/g
4 BLNK-PREP	0	0	01C-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5 SAMPLE	S98T002995	0 W	01C-01 F-02	SOLID	N/A	2.695e+03	97.010	ug/g
5 SAMPLE	S98T002995	0 W	01C-01 NO3-02	SOLID	N/A	3.509e+05	2768.000	ug/g
5 SAMPLE	S98T002995	0 W	01C-01 P04-02	SOLID	N/A	7.485e+03	970.100	ug/g
5 SAMPLE	S98T002995	0 W	01C-01 S04-02	SOLID	N/A	6.649e+03	1116.000	ug/g
5 SAMPLE	S98T002995	0 W	01C-01 OXALATE2	SOLID	N/A	5.527e+03	848.700	ug/g
6 DUP	S98T002995	0 W	01C-01 F-02	SOLID	2.70e+03	6.99e+03	88.545	RPD
6 DUP	S98T002995	0 W	01C-01 CL-02	SOLID	?	1.28e+03		RPD
6 DUP	S98T002995	0 W	01C-01 NO2-02	SOLID	?	1.49e+04		RPD
6 DUP	S98T002995	0 W	01C-01 BR-02	SOLID	?	<1.01e-3		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27292

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002995	0 W	QIC-01	N03-02	SOLID	3.51e+05	1.34e+05	89.485 RPD
6 DUP	S98T002995	0 W	QIC-01	P04-02	SOLID	7.48e+03	2.60e+03	96.825 RPD
6 DUP	S98T002995	0 W	QIC-01	S04-02	SOLID	6.65e+02	2.86e+03	79.706 RPD
6 DUP	S98T002995	0 W	QIC-01	OXALATE2	SOLID	5.53e+03	1.05e+04	62.009 RPD
7 SPK	S98T002995	0 W	QIC-01	F-02	SOLID	5.90e1	6.62e+01	112.203 % Recovery
7 SPK	S98T002995	0 W	QIC-01	CL-02	SOLID	8.00e1	0.00e+00	0.000 % Recovery
7 SPK	S98T002995	0 W	QIC-01	N02-02	SOLID	5.48e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002995	0 W	QIC-01	BR-02	SOLID	5.86e2	0.00e+00	0.000 % Recovery
7 SPK	S98T002995	0 W	QIC-01	N03-02	SOLID	5.92e2	5.86e+02	98.986 % Recovery
7 SPK	S98T002995	0 W	QIC-01	P04-02	SOLID	5.47e2	5.19e+02	94.881 % Recovery
7 SPK	S98T002995	0 W	QIC-01	S04-02	SOLID	6.38e2	6.62e+02	103.762 % Recovery
7 SPK	S98T002995	0 W	QIC-01	OXALATE2	SOLID	5.40e2	5.59e+02	103.519 % Recovery
8 SAMPLE	S98T002997	0 W	QIC-01	N03-02	SOLID	N/A	4.951e+05	2763.000 ug/g
9 DUP	S98T002997	0 W	QIC-01	F-02	SOLID	?	1.57e+03	RPD
9 DUP	S98T002997	0 W	QIC-01	CL-02	SOLID	?	8.76e+02	RPD
9 DUP	S98T002997	0 W	QIC-01	N02-02	SOLID	?	1.07e+04	RPD
9 DUP	S98T002997	0 W	QIC-01	BR-02	SOLID	?	<1.01e3	RPD
9 DUP	S98T002997	0 W	QIC-01	N03-02	SOLID	4.95e+05	5.26e+05	6.072 RPD
9 DUP	S98T002997	0 W	QIC-01	P04-02	SOLID	?	5.10e+03	RPD
9 DUP	S98T002997	0 W	QIC-01	S04-02	SOLID	?	5.05e+03	RPD
9 DUP	S98T002997	0 W	QIC-01	OXALATE2	SOLID	?	2.77e+03	RPD

Final page for worklist# 27292

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

James M. Lyle 12/1/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27292

Analyst: JUM Instrument: IC ✓4052 Book# 24N01B LCS
28N01B CCV

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105(2) @ic-01 rerun. Use 101df for NO3 and 41df for rest.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002995 0 W	@IC-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: F-02			, NO3-02 , OXALATE2, PO4-02 , SO4-02		
6 DUP	S98T002995 0 W	@IC-01	SOLID			
7 SPK	S98T002995 0 W	@IC-01	SOLID			
8 SAMPLE	S98T002997 0 W	@IC-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: NO3-02					
9 DUP	S98T002997 0 W	@IC-01	SOLID			

Final page for worklist # 27292

Valerie L. Mason 11-25-98

Analyst Signature

Date

Analyst Signature

Date

S98T002995 is non homogenous - it has high RPD'S
If better RPD'S are needed a Reprep should
be done.

Validated 12/1/98 Jm Fyfe

Data Entry Comments:

uploaded 12-1-98

John Wamell

27292NOV.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB

Date: 11/25/1998 09:07:44

Data File : C:\DX\DATA\98112501.D01

Method : C:\DX\METHOD\4000SYS2.MET

ACI Address: 1 System: 2 Inject#: 1

Detector: CDM-1

Analyst: *Valerie Mann* 11-25-98 Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.24		0.000	48	403		1
Totals			0.000	48	403		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1335 TO 1348.

Sample Name: 29N21-B LCS Date: 11/25/1998 09:21:22
 Data File : C:\DX\DATA\98112501.D02
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.21		0.000	38	268	1	
2	0.79		0.000	68	271	2	
3	0.95	fluoride	109.4	64.548	1947	5590	2 0.35
4	1.47	chloride	105.09	84.069	1435	4708	1 -2.21
5	1.77	nitrite	97.25	532.930	5384	20793	1 -3.10
6	2.70	bromide	105.10	615.860	3117	15914	1 -3.91
7	3.07	nitrate	106.37	629.727	3531	21670	1 0.11
8	4.45	phosphate	99.43	543.901	864	9294	1 -4.02
9	5.92	sulfate	104.48	666.573	2113	28327	1 -3.48
10	7.79	oxalate	104.31	565.947	1094	19011	1 -3.31
Totals			3703.554		19591	125846	

File: 98112501.D02 Sample: 29N21-B LCS

Sample Name: 28N21-B CCV Date: 11/25/1998 10:10:36
 Data File : C:\DX\DATA\98112501.D03
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	31	171	1	
2	0.79		0.000	72	278	2	
3	0.95	fluoride	101.41 64.908	1942	5622	2	0.35
4	1.47	chloride	105.89 95.305	1573	5350	1	-2.65
5	1.77	nitrite	97.83 549.828	5405	21472	1	-3.10
6	2.69	bromide	108.09 682.240	3339	17683	1	-4.15
7	3.07	nitrate	108.09 754.452	4241	26106	1	-0.11
8	4.45	phosphate	100.23 633.428	993	10858	1	-4.02
9	5.92	sulfate	102.63 717.394	2254	30535	1	-3.48
10	7.79	oxalate	105.52 583.506	1117	19605	1	-3.31
Totals			4081.061	20967	137680		

File: 98112501.D03 Sample: 28N21-B CCV

=====

Sample Name: PREP BLANK	Date: 11/25/1998 13:38:32
Data File : C:\DX\DATA\98112511.D01	
Method : C:\DX\METHOD\400ASYS2.met	
ACI Address: 1 System: 2 Inject#: 1	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.22		0.000	40	341	1	
2	1.46	chloride	0.034	26	172	1	-3.10
3	1.77	nitrite	0.499	456	1668	1	-3.46
4	3.07	nitrate	0.155	49	280	1	0.11

Totals 0.688 571 2461

File: 98112511.D01 Sample: PREP BLANK

Data Reprocessed On 11/30/1998 15:08:20

Sample Name: S98T002995	Date: 11/25/1998 13:50:58
Data File : F:\DATA\98112511.D02	
Method : C:\DX\METHOD\400ASYS2.met	
ACI Address: 1 System: 2 Inject#: 2	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	49	186	1	
2	0.95	fluoride	13.234	409	1108	1	0.35
3	1.46	chloride	6.259	102	330	1	-3.10
4	1.75	nitrite	81.228	739	2875	1	-4.19
5	3.00	nitrate	1779.888	10035	63871	1	7.14
6	4.45	phosphate	43.639	56	682	1	-4.02
7	5.92	sulfate	37.668	91	1464	1	-3.48
8	7.79	oxalate	27.182	54	980	1	-3.31
Totals			1989.097	11534	71498		

File: 98112511.D02 Sample: S98T002995

Data Reprocessed On 11/30/1998 15:09:27

Sample Name: S98T002995	Date: 11/25/1998 14:53:15
Data File : F:\DATA\98112501.D07	
Method : C:\DX\METHOD\400ASYS2.met	
ACI Address: 1 System: 2 Inject#: 7	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	41	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	128	526	1	
2	0.95	fluoride	13.667	965	2881	1	-0.35
3	1.46	chloride	7.903	288	1065	1	-3.10
4	1.76	nitrite	76.182	1828	7059	1	-3.83
5	2.95	nitrate	1738.085	24585	168563	1	5.48
6	4.45	phosphate	37.966	143	1532	1	-4.02
7	5.92	sulfate	33.722	244	3376	1	-3.48
8	7.79	oxalate	28.035	138	2370	1	-3.31
Totals			1935.560	28319	187372		

File: 98112501.D07 Sample: S98T002995

=====

Sample Name: S98T002995 DUP Date: 11/25/1998 14:03:25

Data File : C:\DX\DATA\98112511.D03

Method : C:\DX\METHOD\400ASYS2.met

ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

-----

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	58	222	1	
2	0.95	fluoride	33.469	1004	2864	1	-0.35
3	1.46	chloride	6.849	105	363	1	-3.10
4	1.75	nitrite	79.695	742	2815	1	-4.19
5	3.03	nitrate	677.070	3851	23350	1	-1.41
6	4.48	phosphate	12.965	16	161	1	-3.45
7	5.97	sulfate	15.814	44	545	1	-2.61
8	7.79	oxalate	55.006	103	1903	1	-3.31
Totals			880.867	5924	32223		

File: 98112511.D03 Sample: S98T002995 DUP

=====

Sample Name: S98T002995 DUP	Date: 11/25/1998 15:05:42
Data File : C:\DX\DATA\98112501.D08	
Method : C:\DX\METHOD\400ASYS2.met	
ACI Address: 1 System: 2 Inject#: 8	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	136	555	2	
2	0.95	fluoride	35.347	2566	7589	2	-0.35
3	1.46	chloride	6.476	246	869	1	-3.10
4	1.76	nitrite	75.275	1799	6971	1	-3.83
5	3.01	nitrate	663.095	9337	58329	1	-1.85
6	4.45	phosphate	13.122	46	490	1	-4.02
7	5.92	sulfate	14.479	95	1380	1	-3.48
8	7.79	oxalate	52.986	260	4412	1	-3.31
Totals			860.780	14486	80595		

File: 98112501.D08 Sample: S98T002995 DUP

Sample Name: S98T002995 SPK Date: 11/25/1998 14:15:53
 Data File : C:\DX\DATA\98112501.D04
 Method : C:\DX\METHOD\400ASYS2.met
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	112	451	2	
2	0.95	fluoride	78.125	2353	6793	2	-0.35
3	1.46	chloride	88.319	1500	4950	1	-3.10
4	1.76	nitrite	602.361	6127	23586	1	-3.83
5	2.65	bromide	607.616	3165	15695	1	-5.58
6	2.98	nitrate	2366.073	13680	86619	1	-2.93
7	4.45	phosphate	584.104	895	9996	1	-4.02
8	5.92	sulfate	698.239	2108	29702	1	-3.48
9	7.79	oxalate	571.864	1133	19211	1	-3.31
Totals			5596.701	31073	197003		

File: 98112501.D04 Sample: S98T002995 SPK

Sample Name: S98T002995 SPK Date: 11/25/1998 15:18:09
 Data File : C:\DX\DATA\98112501.D09
 Method : C:\DX\METHOD\400ASYS2.met
 ACI Address: 1 System: 2 Inject#: 9 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	41	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.19		0.000	20	103	1	
2	0.79		0.000	176	676	2	
3	0.95	fluoride	40.146	3013	8646	2	0.35
4	1.46	chloride	40.749	1605	5641	1	-3.10
5	1.76	nitrite	295.245	7182	28639	1	-3.83
6	2.67	bromide	247.461	3147	15747	1	-5.10
7	2.95	nitrate	1995.205	28275	199004	1	-3.80
8	4.45	phosphate	245.738	973	10367	1	-4.02
9	5.92	sulfate	298.485	2308	31314	1	-3.48
10	7.79	oxalate	251.471	1219	20822	1	-3.31
Totals			3414.499	47918	320959		

File: 98112501.D09 Sample: S98T002995 SPK

Sample Name: S98T002997 Date: 11/25/1998 14:28:20
 Data File : F:\DATA\98112501.D05
 Method : C:\DX\METHOD\400ASYS2.met
 ACI Address: 1 System: 2 Inject#: 5 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	52	204	2	
2	0.95	fluoride	8.528	237	702	2	-0.35
3	1.46	chloride	5.127	82	267	1	-3.10
4	1.75	nitrite	62.883	559	2157	1	-4.19
5	2.98	nitrate	2515.063	14411	92551	1	6.43
6	4.45	phosphate	32.525	46	493	1	-4.02
7	5.92	sulfate	27.054	64	1018	1	-3.48
8	7.79	oxalate	14.429	25	558	1	-3.31
Totals			2665.610	15476	97950		

File: 98112501.D05 Sample: S98T002997

Data Reprocessed On 11/30/1998 15:12:00

```

=====
Sample Name: S98T002997                      Date: 11/25/1998 15:30:36
Data File  : F:\DATA\98112501.D10
Method     : C:\DX\METHOD\400ASYS2.met
ACI Address: 1 System: 2 Inject#: 10          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      41    3000    5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	128	530	2	
2	0.95	fluoride	8.183	604	1707	2	0.35
3	1.46	chloride	5.249	192	701	1	-3.10
4	1.76	nitrite	59.822	1387	5472	1	-3.83
5	2.94	nitrate	2435.756	35071	256758	1	5.00
6	4.45	phosphate	32.264	122	1293	1	-4.02
7	5.92	sulfate	23.587	163	2324	1	-3.48
8	7.79	oxalate	10.013	52	897	1	-3.31
Totals			2574.873	37718	269681		

File: 98112501.D10 Sample: S98T002997

Data Reprocessed On 11/30/1998 15:12:34

Sample Name: S98T002997 DUP	Date: 11/25/1998 14:40:47
Data File : F:\DATA\98112501.D06	
Method : C:\DX\METHOD\400ASYS2.met	
ACI Address: 1 System: 2 Inject#: 6	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	50	203	2	
2	0.95	fluoride	8.194	238	673	2	0.35
3	1.46	chloride	4.799	77	249	1	-3.10
4	1.75	nitrite	57.323	499	1939	1	-4.19
5	2.99	nitrate	2662.118	15094	98471	1	6.67
6	4.45	phosphate	26.607	37	393	1	-4.02
7	5.92	sulfate	28.381	72	1073	1	-3.48
8	7.84	oxalate	10.457	28	426	1	-2.65
Totals			2797.879	16095	103427		

File: 98112501.D06 Sample: S98T002997 DUP

Data Reprocessed On 11/30/1998 15:13:10

```

=====
Sample Name: S98T002997 DUP                      Date: 11/25/1998 15:43:04
Data File  : F:\DATA\98112501.D11
Method     : C:\DX\METHOD\400ASYS2.met
ACI Address: 1 System: 2 Inject#: 11              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000   5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	136	579	2	
2	0.95	fluoride	7.951	586	1658	2	0.35
3	1.47	chloride	4.431	175	589	1	-2.65
4	1.77	nitrite	53.912	1274	4900	1	-3.46
5	2.95	nitrate	2517.748	36619	268499	1	5.24
6	4.45	phosphate	25.787	95	1021	1	-4.02
7	5.92	sulfate	25.527	176	2525	1	-3.48
8	7.79	oxalate	14.001	62	1223	1	-3.31
Totals			2649.357	39124	280993		

File: 98112501.D11 Sample: S98T002997 DUP

LABCORE Completed Worklist Report for Worklist# 27395

Analyst: tlm

Instrument: IC40S2

Book#: 29N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1	CCB	0	@IC-QC	F	QC	1	<1.20e-2	ug/mL	
1	CCB	0	@IC-QC	CL	QC	1	<1.70e-2	ug/mL	
1	CCB	0	@IC-QC	NO2	QC	1	<1.08e-1	ug/mL	
1	CCB	0	@IC-QC	BR	QC	1	<1.25e-1	ug/mL	
1	CCB	0	@IC-QC	NO3	QC	1	<1.39e-1	ug/mL	
1	CCB	0	@IC-QC	P04	QC	1	<1.20e-1	ug/mL	
1	CCB	0	@IC-QC	S04	QC	1	<1.38e-1	ug/mL	
1	CCB	0	@IC-QC	OXALATE2	QC	1	<1.05e-1	ug/mL	
2	LCS-INST	0	@IC-QC	F	QC	5.90e1	6.31e+01	106.949 % Recovery	
2	LCS-INST	0	@IC-QC	CL	QC	8.00e1	8.17e+01	102.125 % Recovery	
2	LCS-INST	0	@IC-QC	NO2	QC	5.44e2	5.25e+02	96.507 % Recovery	
2	LCS-INST	0	@IC-QC	BR	QC	5.86e2	5.93e+02	101.195 % Recovery	
2	LCS-INST	0	@IC-QC	NO3	QC	5.92e2	5.81e+02	98.142 % Recovery	
2	LCS-INST	0	@IC-QC	P04	QC	5.47e2	5.41e+02	98.903 % Recovery	
2	LCS-INST	0	@IC-QC	S04	QC	6.38e2	6.40e+02	100.313 % Recovery	
2	LCS-INST	0	@IC-QC	OXALATE2	QC	5.69e2	5.75e+02	101.054 % Recovery	
3	CCV	0	@IC-QC	F	QC	6.40e1	6.32e+01	98.750 % Recovery	
3	CCV	0	@IC-QC	CL	QC	9.00e1	8.97e+01	99.667 % Recovery	
3	CCV	0	@IC-QC	NO2	QC	5.74e2	5.47e+02	95.296 % Recovery	
3	CCV	0	@IC-QC	BR	QC	6.30e2	6.22e+02	98.730 % Recovery	
3	CCV	0	@IC-QC	NO3	QC	6.98e2	6.93e+02	99.284 % Recovery	
3	CCV	0	@IC-QC	P04	QC	6.32e2	6.22e+02	98.418 % Recovery	
3	CCV	0	@IC-QC	S04	QC	6.99e2	6.80e+02	97.282 % Recovery	
3	CCV	0	@IC-QC	OXALATE2	QC	5.71e2	5.68e+02	99.475 % Recovery	
4	BLNK-PREP	0	@IC-01	F-02	SOLID	1	<1.20e-2	ug/g	
4	BLNK-PREP	0	@IC-01	CL-02	SOLID	1	<1.70e-2	ug/g	
4	BLNK-PREP	0	@IC-01	NO2-02	SOLID	1	7.30e-01	0.730 ug/g	
4	BLNK-PREP	0	@IC-01	BR-02	SOLID	1	<1.25e-1	ug/g	
4	BLNK-PREP	0	@IC-01	NO3-02	SOLID	1	3.05e-01	0.305 ug/g	
4	BLNK-PREP	0	@IC-01	P04-02	SOLID	1	<1.20e-1	ug/g	
4	BLNK-PREP	0	@IC-01	S04-02	SOLID	1	<1.38e-1	ug/g	
4	BLNK-PREP	0	@IC-01	OXALATE2	SOLID	1	<1.05e-1	ug/g	
5	SAMPLE	S98T002610	0 W	@IC-01	F-02	SOLID	N/A	< 2.285e+02	228.500 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	CL-02	SOLID	N/A	< 3.237e+02	323.700 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	NO2-02	SOLID	N/A	< 3.165e+03	2056.000 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	BR-02	SOLID	N/A	< 2.380e+03	2381.000 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	NO3-02	SOLID	N/A	< 6.898e+05	2647.000 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	P04-02	SOLID	N/A	< 2.285e+03	2285.000 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	S04-02	SOLID	N/A	< 2.628e+03	2628.000 ug/g
5	SAMPLE	S98T002610	0 W	@IC-01	OXALATE2	SOLID	N/A	< 1.999e+03	2000.000 ug/g
6	DUP	S98T002610	0 W	@IC-01	F-02	SOLID	<2.28e2	<2.27e2	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27395

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002610	0 W	@IC-01	CL-02	SOLID	<3.24e2	<3.22e2	RPD
6 DUP	S98T002610	0 W	@IC-01	NO2-02	SOLID	3.16e+03	3.25e+03	2.808 RPD
6 DUP	S98T002610	0 W	@IC-01	BR-02	SOLID	<2.38e3	<2.37e3	RPD
6 DUP	S98T002610	0 W	@IC-01	NO3-02	SOLID	6.90e+05	6.49e+05	6.124 RPD
6 DUP	S98T002610	0 W	@IC-01	PO4-02	SOLID	<2.28e3	<2.27e3	RPD
6 DUP	S98T002610	0 W	@IC-01	SO4-02	SOLID	<2.63e3	<2.62e3	RPD
6 DUP	S98T002610	0 W	@IC-01	OXALATE2	SOLID	<2.00e3	<1.99e3	RPD
7 SPK	S98T002610	0 W	@IC-01	F-02	SOLID	5.90e1	6.15e+01	104.237 % Recovery
7 SPK	S98T002610	0 W	@IC-01	CL-02	SOLID	8.00e1	7.90e+01	98.750 % Recovery
7 SPK	S98T002610	0 W	@IC-01	NO2-02	SOLID	5.44e2	4.95e+02	90.993 % Recovery
7 SPK	S98T002610	0 W	@IC-01	BR-02	SOLID	5.86e2	5.69e+02	97.059 % Recovery
7 SPK	S98T002610	0 W	@IC-01	NO3-02	SOLID	5.92e2	5.52e+02	93.243 % Recovery
7 SPK	S98T002610	0 W	@IC-01	PO4-02	SOLID	5.47e2	5.19e+02	94.981 % Recovery
7 SPK	S98T002610	0 W	@IC-01	SO4-02	SOLID	6.38e2	6.20e+02	97.179 % Recovery
7 SPK	S98T002610	0 W	@IC-01	OXALATE2	SOLID	5.69e2	5.50e+02	96.661 % Recovery
8 SAMPLE	S98T002611	0 W	@IC-01	F-02	SOLID	N/A	4.268e+02	232.300 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	CL-02	SOLID	N/A	3.291e+02	329.100 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	NO2-02	SOLID	N/A	5.198e+03	2091.000 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	BR-02	SOLID	N/A	2.420e+03	2420.000 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	NO3-02	SOLID	N/A	6.161e+05	2691.000 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	PO4-02	SOLID	N/A	2.323e+03	2323.000 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	SO4-02	SOLID	N/A	2.671e+03	2671.000 ug/g
8 SAMPLE	S98T002611	0 W	@IC-01	OXALATE2	SOLID	N/A	2.032e+03	2032.000 ug/g
9 DUP	S98T002611	0 W	@IC-01	F-02	SOLID	4.27e+02	4.36e+02	2.086 RPD
9 DUP	S98T002611	0 W	@IC-01	CL-02	SOLID	<3.29e2	<3.30e2	RPD
9 DUP	S98T002611	0 W	@IC-01	NO2-02	SOLID	5.20e+03	5.70e+03	9.174 RPD
9 DUP	S98T002611	0 W	@IC-01	BR-02	SOLID	<2.42e3	<2.42e3	RPD
9 DUP	S98T002611	0 W	@IC-01	NO3-02	SOLID	6.16e+05	5.73e+05	7.233 RPD
9 DUP	S98T002611	0 W	@IC-01	PO4-02	SOLID	<2.32e3	<2.33e3	RPD
9 DUP	S98T002611	0 W	@IC-01	SO4-02	SOLID	<2.67e3	<2.68e3	RPD
9 DUP	S98T002611	0 W	@IC-01	OXALATE2	SOLID	<2.03e3	<2.04e3	RPD
10 SAMPLE	S98T002612	0 W	@IC-01	F-02	SOLID	N/A	7.266e+03	231.300 ug/g
10 SAMPLE	S98T002612	0 W	@IC-01	CL-02	SOLID	N/A	3.276e+02	327.600 ug/g
10 SAMPLE	S98T002612	0 W	@IC-01	PO4-02	SOLID	N/A	1.741e+04	2313.000 ug/g
10 SAMPLE	S98T002612	0 W	@IC-01	OXALATE2	SOLID	N/A	1.233e+04	2023.000 ug/g
11 DUP	S98T002612	0 W	@IC-01	F-02	SOLID	7.27e+03	8.58e+03	16.530 RPD
11 DUP	S98T002612	0 W	@IC-01	CL-02	SOLID	<3.28e2	<3.39e2	RPD
11 DUP	S98T002612	0 W	@IC-01	NO2-02	SOLID	6.82e+03	8.68e+03	24.000 RPD
11 DUP	S98T002612	0 W	@IC-01	BR-02	SOLID	<2.41e3	<2.41e3	RPD
11 DUP	S98T002612	0 W	@IC-01	NO3-02	SOLID	2.94e+05	1.75e+05	50.746 RPD
11 DUP	S98T002612	0 W	@IC-01	PO4-02	SOLID	1.74e+04	1.63e+04	6.528 RPD
11 DUP	S98T002612	0 W	@IC-01	SO4-02	SOLID	1.38e+04	1.70e+04	20.779 RPD
11 DUP	S98T002612	0 W	@IC-01	OXALATE2	SOLID	1.23e+04	1.41e+04	13.636 RPD

Final page for worklist# 27395

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27395

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 27395

Analyst: znt Instrument: IC 4052 Book# CCV 28N21C
Method: LA-533-105 Rev/Mod F-0 LCS 29N21C
Worklist Comment: BY105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	CCB		@IC-QC	QC		
2	LCS-INST		@IC-QC	QC		
3	CCV		@IC-QC	QC		
4	BLNK-PREP		@IC-01	SOLID		
5	SAMPLE	S98T002610 0 W	@IC-01	SOLID	98000436	BY-105 (2)
	Analyses Requested:	BR-02		CL-02	F-02	NO2-02 , NO3-02 ,
		OKALATE2,	PO4-02		SO4-02	
6	DUP	S98T002610 0 W	@IC-01	SOLID		
7	SPK	S98T002610 0 W	@IC-01	SOLID		
8	SAMPLE	S98T002611 0 W	@IC-01	SOLID	98000436	BY-105 (2)
	Analyses Requested:	BR-02		CL-02	F-02	NO2-02 , NO3-02 ,
		OKALATE2,	PO4-02		SO4-02	
9	DUP	S98T002611 0 W	@IC-01	SOLID		
10	SAMPLE	S98T002612 0 W	@IC-01	SOLID	98000436	BY-105 (2)
	Analyses Requested:	BR-02		CL-02	F-02	NO2-02 , NO3-02 ,
		OKALATE2,	PO4-02		SO4-02	
11	DUP	S98T002612 0 W	@IC-01	SOLID		

Final page for worklist # 27395

Seena Lee Mui 12-10-98
Analyst Signature Date

Analyst Signature Date

S98T002612 NO₂, Br, NO₃, SO₄
rejected for high RPD & too < det.
limit.

Data Entry Comments:
uploaded 12-10-98

W. Wallace
27395 DEC.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Rest of samples validated even
though Br < higher than desired
NO₃ level too high to allow
Br to be run at desired level.

S98T002612 reanalyzed on
W# 27732. FS 3/24/99

AM Luy
12/18/98

Sample Name: CCB Date: 12/10/1998 09:28:01
Data File : C:\DX\DATA\98121051.D01
Method : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-1
Analyst : *LM* Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta

Totals			400.000	0	0		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1353 TO 1363.

Sample Name: ~~CCV-29N21C~~ **CCV-29N21C LCS** Date: 12/10/1998 10:07:43
 Data File : C:\DX\DATA\98121051.D03
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-1
 Analyst : *LM* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	64	249	2	
2	0.95	fluoride	63.091	1906	5669	2	-0.35
3	1.47	chloride	81.715	1435	4807	1	-1.35
4	1.77	nitrite	524.821	5339	20668	1	-1.48
5	2.71	bromide	593.492	3153	16207	1	-2.29
6	3.08	nitrate	580.723	3531	21827	1	-2.84
7	4.40	phosphate	541.151	865	9459	1	0.00
8	5.81	sulfate	640.220	2172	28860	1	-0.97
9	7.63	oxalate	575.476	1133	20539	1	-1.34

Totals 3600.690 19598 128285

File: 98121051.D03 Sample: CCV 29N21C

Sample Name: CCV 28N21C Date: 12/10/1998 09:50:33
 Data File : C:\DX\DATA\98121051.D02
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-1
 Analyst : *lm* Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration $\mu\text{g/ml}$	Height	Area	Bl. Code	%Delta
1	0.80		0.000	65	244	2	
2	0.95	fluoride	63.243	1955	5683	2	0.35
3	1.47	chloride	89.680	1584	5278	1	-0.90
4	1.77	nitrite	546.622	5505	21553	1	-1.48
5	2.71	bromide	622.476	3300	17024	1	-2.05
6	3.09	nitrate	693.120	4205	26222	1	-2.63
7	4.40	phosphate	621.576	989	10892	1	0.00
8	5.81	sulfate	679.724	2240	30685	1	-0.97
9	7.63	oxalate	567.576	1090	20254	1	-1.34

Totals 3884.017 20931 137835

File: 98121051.D02 Sample: CCV 28N21C

Sample Name: BLNK-PREP Date: 12/10/1998 10:21:14
 Data File : C:\DK\DATA\98121051.D04
 Method : C:\DK\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-1
 Analyst : *LM* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.47	chloride	0.003	23	56	1	-0.90
2	1.77	nitrite	0.730	668	2509	1	-1.48
3	3.11	nitrate	0.305	102	621	1	-2.00
Totals			1.032	793	3186		

File: 98121051.D04 Sample: BLNK-PREP

Data Reprocessed On 12/10/1998 15:34:36

```

=====
Sample Name: S98T002610                      Date: 12/10/1998 11:29:56
Data File  : F:\DATA\98121051.D07
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 7           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          101    3000   5Hz   0.00  10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	65	287	1	
2	1.77	nitrite	16.787	62	236	1	-1.85
3	3.00	nitrate	3658.595	22122	151011	1	0.00
Totals			3675.382	22249	151535		

File: 98121051.D07 Sample: S98T002610

Data Reprocessed On 12/10/1998 15:35:02

```

=====
Sample Name: S98T002610 DUP                      Date: 12/10/1998 13:49:14
Data File  : F:\DATA\98121051.D08
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 8                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00         30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	58	230	1	
2	1.77	nitrite	17.319	72	258	1	-1.85
3	3.01	nitrate	3457.590	20794	141933	1	-0.11
Totals			3474.908	20924	142420		

File: 98121051.D08 Sample: S98T002610 DUP

Data Reprocessed On 12/10/1998 15:35:33

```

=====
Sample Name: S98T002610 SPK                      Date: 12/10/1998 14:03:05
Data File  : F:\DATA\98121051.D09
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                      Column: AG4A/AS4A anion column
Detector: CDM-1
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	121	482	2	
2	0.95	fluoride	61.530	1831	5526	2	-0.35
3	1.47	chloride	78.972	1393	4645	1	-1.35
4	1.77	nitrite	511.664	5159	20135	1	-1.85
5	2.69	bromide	569.385	3140	15528	1	-2.77
6	2.99	nitrate	4210.756	24995	176508	1	-0.11
7	4.37	phosphate	519.307	815	9071	1	-0.61
8	5.81	sulfate	620.014	2201	27927	1	-0.97
9	7.63	oxalate	549.595	1147	19606	1	-1.34
Totals			7121.223	40802	279428		

File: 98121051.D09 Sample: S98T002610 SPK

Data Reprocessed On 12/10/1998 15:36:00

```

=====
Sample Name: S98T002611                      Date: 12/10/1998 14:14:41
Data File  : F:\DATA\98121051.D10
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 10          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00  10.00          30
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	59	229	2	
2	0.95	fluoride	2.227	48	130	2	0.35
3	1.47	chloride	0.141	25	79	1	-1.35
4	1.77	nitrite	27.123	175	649	1	-1.85
5	3.00	nitrate	3214.661	19125	131098	1	0.00
Totals			3244.151	19432	132185		

File: 98121051.D10 Sample: S98T002611

Data Reprocessed On 12/10/1998 15:36:30

```

=====
Sample Name: S98T002611 DUP                      Date: 12/10/1998 14:28:41
Data File  : F:\DATA\98121051.D11
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 11              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00         30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	53	213	2	
2	0.95	fluoride	2.269	51	134	2	0.35
3	1.47	chloride	-0.065	23	67	1	-1.35
4	1.77	nitrite	29.679	194	751	1	-1.85
5	3.01	nitrate	2985.255	18209	120998	1	-0.11
Totals			3017.139	18530	122162		

File: 98121051.D11 Sample: S98T002611 DUP

Data Reprocessed On 12/10/1998 15:36:59

```

=====
Sample Name: S98T002612                      Date: 12/10/1998 14:42:07
Data File  : F:\DATA\98121051.D12
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 12          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101      3000      5Hz      0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	56	201	2	
2	0.95	fluoride	38.080	1114	3380	2	-0.35
3	1.47	chloride	1.362	48	149	1	-1.35
4	1.77	nitrite	35.726	259	992	1	-1.85
5	3.04	nitrate	1539.705	9431	60031	1	0.00
6	4.40	phosphate	91.266	137	1524	1	0.00
7	5.81	sulfate	72.370	221	2928	1	-0.97
8	7.63	oxalate	64.641	123	2217	1	-1.34
Totals			1843.150	11390	71422		

File: 98121051.D12 Sample: S98T002612

Sample Name: S98T002612 DUP Date: 12/10/1998 14:54:05
 Data File : C:\DX\DATA\98121051.D13
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 13 Detector: CDM-1
 Analyst : *Lmr* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	56	203	2	
2	0.95	fluoride	44.856	1306	3998	2	0.35
3	1.47	chloride	1.773	54	172	1	-1.35
4	1.77	nitrite	45.367	360	1377	1	-1.85
5	3.06	nitrate	914.547	5451	34942	1	-3.47
6	4.37	phosphate	85.067	124	1416	1	-0.61
7	5.81	sulfate	88.958	282	3677	1	-0.97
8	7.63	oxalate	73.516	143	2534	1	-1.34
Totals			1254.085	7777	48319		

File: 98121051.D13 Sample: S98T002612 DUP

LABCORE Completed Worklist Report for Worklist# 27396

Analyst: adp

Instrument: IC45S2

Book#: 35N21A

Method: LA-533-105

Rev/Mod F-0

Worklist Comment: BY105 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC4G-QC F*4 QC	1	<1.10e-2		ug/mL
1	CCB	0	@IC4G-QC ACETATE2 QC	1	<2.00e-2		ug/mL
1	CCB	0	@IC4G-QC FORMATE2 QC	1	<2.10e-2		ug/mL
1	CCB	0	@IC4G-QC GLYCOLTI QC	1	<2.00e-2		ug/mL
2	LCS-INST	0	@IC4G-QC F*4 QC	6.05e1	6.33e+01	104.628 % Recovery	
2	LCS-INST	0	@IC4G-QC ACETATE2 QC	1.19e2	9.47e+01	79.580 % Recovery	
2	LCS-INST	0	@IC4G-QC FORMATE2 QC	1.28e2	1.20e+02	93.750 % Recovery	
2	LCS-INST	0	@IC4G-QC GLYCOLTI QC	1.16e2	1.10e+02	94.828 % Recovery	
3	CCV	0	@IC4G-QC F*4 QC	6.21e1	6.22e+01	100.161 % Recovery	
3	CCV	0	@IC4G-QC ACETATE2 QC	1.23e2	1.23e+02	100.000 % Recovery	
3	CCV	0	@IC4G-QC FORMATE2 QC	1.30e2	1.29e+02	99.231 % Recovery	
3	CCV	0	@IC4G-QC GLYCOLTI QC	1.02e2	1.01e+02	99.020 % Recovery	
4	BLNK-PREP	0	@IC4G-01 F*4-01 SOLID	1	<3.10e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 ACETATE2 SOLID	1	<2.00e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 FORMATE2 SOLID	1	<6.20e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 GLYCOLTI SOLID	1	<5.60e-2		ug/g
5	SAMPLE	S98T002610	0 W @IC4G-01 F*4-01 SOLID	N/A	7.805e+01	35.070	ug/g
5	SAMPLE	S98T002610	0 W @IC4G-01 ACETATE2 SOLID	N/A	< 2.262e+01	22.630	ug/g
5	SAMPLE	S98T002610	0 W @IC4G-01 FORMATE2 SOLID	N/A	< 7.014e+01	70.140	ug/g
5	SAMPLE	S98T002610	0 W @IC4G-01 GLYCOLTI SOLID	N/A	< 6.335e+01	63.360	ug/g
6	DUP	S98T002610	0 W @IC4G-01 F*4-01 SOLID	7.80e+01	9.03e+01	14.617	RPD
6	DUP	S98T002610	0 W @IC4G-01 ACETATE2 SOLID	<2.26e1	2.83e+01		RPD
6	DUP	S98T002610	0 W @IC4G-01 FORMATE2 SOLID	<7.01e1	<6.36e1		RPD
6	DUP	S98T002610	0 W @IC4G-01 GLYCOLTI SOLID	<6.34e1	<6.31e1		RPD
7	SPK	S98T002610	0 W @IC4G-01 F*4-01 SOLID	6.05e1	6.06e+01	100.165 % Recovery	
7	SPK	S98T002610	0 W @IC4G-01 ACETATE2 SOLID	1.19e2	9.46e+01	79.496 % Recovery	
7	SPK	S98T002610	0 W @IC4G-01 FORMATE2 SOLID	1.28e2	1.19e+02	92.969 % Recovery	
7	SPK	S98T002610	0 W @IC4G-01 GLYCOLTI SOLID	1.16e2	1.05e+02	90.517 % Recovery	
8	SAMPLE	S98T002611	0 W @IC4G-01 F*4-01 SOLID	N/A	2.936e+02	35.650	ug/g
8	SAMPLE	S98T002611	0 W @IC4G-01 ACETATE2 SOLID	N/A	6.094e+01	23.000	ug/g
8	SAMPLE	S98T002611	0 W @IC4G-01 FORMATE2 SOLID	N/A	1.140e+02	71.280	ug/g
8	SAMPLE	S98T002611	0 W @IC4G-01 GLYCOLTI SOLID	N/A	< 6.439e+01	64.380	ug/g
9	DUP	S98T002611	0 W @IC4G-01 F*4-01 SOLID	2.94e+02	3.12e+02	5.941	RPD
9	DUP	S98T002611	0 W @IC4G-01 ACETATE2 SOLID	6.09e+01	5.74e+01	5.917	RPD
9	DUP	S98T002611	0 W @IC4G-01 FORMATE2 SOLID	1.14e+02	1.21e+02	5.957	RPD
9	DUP	S98T002611	0 W @IC4G-01 GLYCOLTI SOLID	<6.44e1	<6.45e1		RPD
10	SAMPLE	S98T002612	0 W @IC4G-01 F*4-01 SOLID	N/A	7.018e+03	35.490	ug/g
10	SAMPLE	S98T002612	0 W @IC4G-01 GLYCOLTI SOLID	N/A	< 6.411e+01	64.080	ug/g
11	DUP	S98T002612	0 W @IC4G-01 F*4-01 SOLID	7.02e+03	7.73e+03	9.627	RPD
11	DUP	S98T002612	0 W @IC4G-01 ACETATE2 SOLID	1.56e+02	2.54e+02	47.805	RPD
11	DUP	S98T002612	0 W @IC4G-01 FORMATE2 SOLID	1.43e+02	2.07e+02	36.571	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27396

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	DUP	S98T002612	0	W	@IC4G-01	GLYCOLTL SOLID	<6.41e1	<6.43e1		RED

Final page for worklist# 27396

_____
Analyst Signature   Date

_____
Analyst Signature   Date

_____
Reviewer Signature   Date 1/25/99

LABCORE Data Entry Template for Worklist# 27396

Analyst: ADD Instrument: IC 4552 Book# 35N21-A LCS
Method: LA-533-105 Rev/Mod E-0 34N21-A CCV
Worklist Comment: BY105 FOR @IC4G-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	LCS-INST				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002610	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
6	DUP	S98T002610	0	W	@IC4G-01	SOLID		
7	SPK	S98T002610	0	W	@IC4G-01	SOLID		
8	SAMPLE	S98T002611	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
9	DUP	S98T002611	0	W	@IC4G-01	SOLID		
10	SAMPLE	S98T002612	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
11	DUP	S98T002612	0	W	@IC4G-01	SOLID		

Final page for worklist # 27396

[Signature] 01-22-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 1-22-99
JH Warrall
27396 JAN.CSV

S98T002612 acetate + formate
rejected for high RPD, Rest
Validated 1/25/99 gm Juy

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

S98T002612 reanalyzed on WL# 28251. Star 3/24/99

Sample Name: 35N21-A LCS Date: 01/22/1999 00:41:01
 Data File : C:\DX\DATA\99012101.D11
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 11 Detector: CDM-2
 Analyst : *Anthony Karamba* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	5102	249179	3	
3	2.13		0.000	31173	522024	2	
4	2.29		0.000	30692	356163	2	
5	2.75		0.000	28785	667776	2	
6	3.00		0.000	26698	723293	2	
7	3.43		0.000	19803	353786	2	
8	3.72		0.000	15896	78479	2	
9	3.80		0.000	16616	154065	2	
10	3.99		0.000	13942	141183	2	
11	4.25		0.000	10569	156932	2	
12	4.43		0.000	7521	51822	2	
13	4.67		0.000	4941	68517	2	
23	7.12	fluoride	63.323 ^{104.7}	1893285	27205204	2	0.00
24	7.79	glycolate	110.480 ^{95.2}	545207	9840870	2	0.00
25	8.27	acetate	94.665 ^{79.6}	501439	9520319	2	0.00
26	8.96	formate	119.505 ^{93.4}	1067275	20084587	2	0.00
Totals			387.972	4218944	70174198		

File: 99012101.D11 Sample: 35N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1367 TO 1377.


```

=====
Sample Name: 34N21-A CCV                               Date: 01/22/1999 01:04:15
Data File  : C:\DX\DATA\99012101.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12                    Detector: CDM-2
Analyst    :                                           Column: AG14A-SC,AS14A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz    0.00 19.00    50000
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	3999	86801	2	
2	1.03		0.000	8954	344160	2	
25	7.12	fluoride	62.174	1867718	26697125	2	0.00
26	7.79	glycolate	101.066	509927	8997548	2	0.00
27	8.27	acetate	123.202	641989	12311726	2	0.00
28	8.96	formate	129.498	1182290	21797181	2	0.00
Totals			415.940	4214877	70234542		

File: 99012101.D12 Sample: 34N21-A CCV

```

=====
Sample Name: BLANK                      Date: 01/22/1999 00:18:26
Data File  : C:\DX\DATA\99012101.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10    Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
5	2.73		0.000	5097	80007	2	
6	2.88		0.000	5311	66446	2	
11	3.85		0.000	3047	80350	2	
26	7.09	fluoride	0.010	4030	88289	2	0.00
30	8.24	acetate	-0.013	4838	93885	2	0.00
Totals			-0.003	22322	408977		

File: 99012101.D10 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 01/22/1999 01:24:27
Data File  : C:\DX\DATA\99012101.D13
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1   System: 2   Inject#: 13      Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume  Dilution Points Rate  Start  Stop Area Reject
-----
External           1           1  5700  5Hz   0.00  19.00   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
7	2.88		0.000	3960	61007	2	
8	3.09		0.000	3931	54432	2	
33	7.39	fluoride	0.010	3369	67369	2	0.00
35	8.29	acetate	-0.013	3564	96366	2	0.00
37	9.41		0.000	3288	82894	2	
Totals			-0.003	18111	362068		

File: 99012101.D13 Sample: PREP BLANK

Data Reprocessed On 01/22/1999 13:46:20

```

=====
Sample Name: S98T002610 SAM                      Date: 01/22/1999 01:45:13
Data File  : E:\DATA\99012101.D14
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1           6    5700   5Hz   0.00 19.00    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	141812	1819874	1	
4	2.95		0.000	14585	299742	2	
5	3.04		0.000	13505	181978	2	
6	3.33		0.000	9329	136280	2	
17	5.35		0.000	5148	71607	2	
22	6.99	fluoride	0.414	145205	2639340	3	0.00
23	7.65	glycolate	-0.066	3751	74583	4	0.00
24	8.13	acetate	0.059	16473	321392	2	0.00
25	8.85	formate	0.213	23608	548009	2	0.00
26	9.63		0.000	5339	95132	2	
Totals			0.621	378757	6187935		

File: 99012101.D14 Sample: S98T002610 SAM

Data Reprocessed On 01/22/1999 13:48:51

```

=====
Sample Name: S98T002610  DUP                      Date: 01/22/1999 02:05:02
Data File  : E:\DATA\99012101.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 15              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   0.00 19.00   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.67		0.000	146139	1962098	1	
4	2.73		0.000	4365	63620	2	
7	3.08		0.000	3364	50153	2	
24	6.96	fluoride	0.481	173577	3120412	3	0.00
27	8.11	acetate	0.151	20385	474172	2	0.00
28	8.83	formate	0.361	45338	961120	2	0.00
Totals			0.993	393168	6631575		

File: 99012101.D15 Sample: S98T002610 DUP

Data Reprocessed On 01/22/1999 13:50:01

```

=====
Sample Name: S98T002610 SPK                      Date: 01/22/1999 02:25:30
Data File  : E:\DATA\99012101.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====
    
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   0.00 19.00   50000
    
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	4490	151242	1	
2	1.70		0.000	150952	1904418	1	
3	2.88		0.000	14545	246842	2	
4	2.95		0.000	14633	72333	2	
5	3.05		0.000	13494	254333	2	
6	3.52		0.000	5369	59403	2	
7	3.80		0.000	4658	58239	2	
20	6.99	fluoride	3.443	1474621	24835906	2	0.00
21	7.68	glycolate	5.274	407916	7903227	2	0.00
22	8.16	acetate	4.731	402144	8048104	2	0.00
23	8.85	formate	5.932	841939	16730128	2	0.00
Totals			19.380	3334760	60264175		

File: 99012101.D16 Sample: S98T002610 SPK

Data Reprocessed On 01/22/1999 13:50:53

```

=====
Sample Name: S98T002611 SAM                      Date: 01/22/1999 02:45:43
Data File  : E:\DATA\99012101.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6  5700  5Hz  0.00 19.00   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.67		0.000	136203	1760317	2	
4	2.93		0.000	10145	306666	2	
5	3.20		0.000	4682	73891	3	
25	6.96	fluoride	1.532	640740	10762379	2	0.00
27	8.13	acetate	0.318	37847	749531	2	0.00
28	8.83	formate	0.595	79545	1615211	2	0.00
29	9.60		0.000	10495	225504	2	
Totals			2.444	919658	15493499		

File: 99012101.D17 Sample: S98T002611 SAM

Data Reprocessed On 01/22/1999 13:51:41

```

=====
Sample Name: S98T002611 DUP                      Date: 01/22/1999 03:05:20
Data File  : E:\DATA\99012101.D18
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1   System: 2   Inject#: 18           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6    5700   5Hz   0.00 19.00    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	2640	76132	1	
2	1.70		0.000	139730	1759658	1	
3	2.92		0.000	12897	230854	2	
4	3.09		0.000	10530	191286	2	
10	4.03		0.000	4443	66270	2	
23	6.99	fluoride	1.625	687232	11443946	3	0.00
25	8.13	acetate	0.299	37967	718056	2	0.00
26	8.85	formate	0.632	86489	1719726	2	0.00
27	9.65		0.000	13162	268019	2	
Totals			2.556	995090	16473947		

File: 99012101.D18 Sample: S98T002611 DUP

Data Reprocessed On 01/22/1999 13:52:48

```

=====
Sample Name: S98T002612 SAM                      Date: 01/22/1999 03:25:10
Data File  : E:\DATA\99012101.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6  5700  5Hz  0.00 19.00   50000
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	147605	1819380	1	
4	3.03		0.000	18142	551348	2	
5	3.27		0.000	14044	91232	2	
6	3.45		0.000	11736	174926	2	
7	3.65		0.000	7251	90575	2	
16	5.24		0.000	3863	52851	2	
21	7.01	fluoride	36.782	18753047	329463598	3	0.00
22	8.13	acetate	0.817	73965	1577003	4	0.00
23	8.85	formate	0.751	101472	2053175	2	0.00
24	9.65		0.000	20157	375860	2	
Totals			38.350	19151280	336249948		

File: 99012101.D19 Sample: S98T002612 SAM

Data Reprocessed On 01/22/1999 13:54:20

```

=====
Sample Name: S98T002612  DUP                      Date: 01/22/1999 03:45:22
Data File  : E:\DATA\99012101.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   0.00 19.00   50000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.67		0.000	145427	1945568	1	
4	2.92		0.000	16312	606246	2	
21	6.99	fluoride	40.393	21030859	374298917	2	0.00
22	8.11	acetate	1.328	104862	2423502	2	0.00
23	8.80	formate	1.083	143401	2982852	2	0.00
24	9.57		0.000	29544	656655	2	
Totals			42.804	21470406	382913740		

File: 99012101.D20 Sample: S98T002612 DUP

LABCORE Completed Worklist Report for Worklist# 27635

Analyst: adp

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 @IC4G-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	5.89e1	6.36e+01	107.980 %	Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.24e2	1.18e+02	95.161 %	Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.23e+02	103.361 %	Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.07e2	1.10e+02	102.804 %	Recovery
3 CCV	0	@IC4G-QC F*4	QC	6.21e1	5.98e+01	96.296 %	Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.22e+02	99.187 %	Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.29e+02	99.231 %	Recovery
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.02e2	1.02e+02	100.000 %	Recovery
4 SAMPLE	S98T003384	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 2.885e+02	288.500	ug/mL
5 DUP	S98T003384	@IC4G-01 F*4-01	LIQUID	2.12e+02	3.15e+02	39.089	RPD
5 DUP	S98T003384	@IC4G-01 ACETATE2	LIQUID	1.36e+03	2.37e+03	54.155	RPD
5 DUP	S98T003384	@IC4G-01 FORMATE2	LIQUID	1.05e+03	1.64e+03	43.866	RPD
5 DUP	S98T003384	@IC4G-01 GLYCOLT1	LIQUID	<2.88e2	<2.88e2		RPD
6 SPK	S98T003384	@IC4G-01 F*4-01	LIQUID	5.89e1	6.39e+01	108.489 %	Recovery
6 SPK	S98T003384	@IC4G-01 ACETATE2	LIQUID	1.24e2	1.47e+02	118.549 %	Recovery
6 SPK	S98T003384	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.37e+02	115.126 %	Recovery
6 SPK	S98T003384	@IC4G-01 GLYCOLT1	LIQUID	1.07e2	1.12e+02	104.673 %	Recovery

Final page for worklist# 27635

Analyst Signature

Date

Analyst Signature

Date

James M. Lyle 12/30/98
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27635

Analyst: ADP Instrument: IC 4552 Book# 35N21-C LCS
Method: LA-533-105 Rev/Mod F-0 34N21-A CCV
Worklist Comment: BY-105 @IC4G-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T003384 0		@IC4G-01 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: ACETATE2, F+4-01, FORMATE2, GLYCOLT1					
5 DUP	S98T003384 0		@IC4G-01 LIQUID			
6 SPK	S98T003384 0		@IC4G-01 LIQUID			

Final page for worklist # 27635

John W. Howell 12-23-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

inputted 12-29-98
John W. Howell
27635 DEC.CSV

acetate, fluoride & formate
reported due to high
RPD

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Reanalyzed on WL# 27751. 12/30/98 JmL
3/24/99

Sample Name: 35N21-C LCS Date: 12/23/1998 03:47:08
 Data File : C:\DX\DATA\98122201.D11
 Method : C:\DX\METHOD\AAS14GLY.MET
 ACI Address: System 2 Inject#: 11 Detector: CDM-2
 Analyst : *Anthony Purita* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 500

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	22400	107800	1	
2	1.60		0.000	5808	37760	1	
3	2.31		0.000	3709	46160	1	
4	2.57		0.000	2059	6857	2	
5	2.75		0.000	5652	49679	2	
6	2.88		0.000	6770	40765	2	
7	2.99		0.000	5305	42592	3	
8	3.23		0.000	340	1581	4	
9	3.31		0.000	508	816	2	
10	3.37		0.000	2494	20070	2	
11	3.60		0.000	3225	16821	2	
12	3.99		0.000	287	3960	1	
13	4.07		0.000	244	1858	2	
14	4.17		0.000	2095	6702	2	
15	4.52		0.000	2963	17409	2	
16	4.72		0.000	3065	26923	2	
17	5.11		0.000	4123	27415	2	
18	5.25		0.000	30231	278157	2	
19	5.72		0.000	2504	33390	2	
20	6.07		0.000	1035	6089	2	
22	6.31		0.000	2958	53816	2	
23	7.39	fluoride	63.607/107.99	1853793	27331059	2	0.00
24	8.11	glycolate	110.499/132.7	531217	9842612	2	0.00
25	8.61	acetate	118.372/95.46	606219	11839668	2	0.00
26	9.33	formate	123.492/123.72	1019642	20767268	2	0.00
27	11.28		0.000	7573491	302647353	2	
28	11.68		0.000	5514282	70840752	2	
29	12.00		0.000	6660915	516306021	3	
30	18.37		0.000	454	61098	4	
Totals			415.970	23867789	960462451		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1380 TO 1391.

```

=====
Sample Name: 34N21-A CCV                      Date: 12/23/1998 04:12:00
Data File  : C:\DX\DATA\98122201.D12
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 500

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.13		0.000	19660	175089	3	
2	0.43		0.000	1796	8203	4	
3	1.30		0.000	3600	68880	1	
4	2.35		0.000	2800	14200	1	
5	2.52		0.000	892	1880	1	
6	2.59		0.000	2560	6020	1	
7	2.67		0.000	2309	10720	1	
9	2.96		0.000	3916	18536	2	
10	3.04		0.000	4809	21798	2	
11	3.12		0.000	5501	28852	2	
12	3.24		0.000	2041	12770	2	
13	3.36		0.000	1980	8318	2	
14	3.51		0.000	4750	39188	2	
15	3.68		0.000	3951	29247	2	
16	3.84		0.000	3736	22944	2	
17	3.96		0.000	4076	75745	2	
18	4.29		0.000	5462	26620	2	
19	4.44		0.000	3632	44983	2	
20	4.64		0.000	6264	54494	2	
21	4.84		0.000	7496	76478	2	
22	5.04		0.000	5328	42779	2	
23	5.25		0.000	40775	360479	3	
24	5.59		0.000	3787	27457	4	
25	5.72		0.000	6147	29326	4	
26	5.81		0.000	8425	145212	2	
27	6.16		0.000	6427	29217	2	
28	6.23		0.000	6704	31720	2	
29	6.37		0.000	9074	85648	2	
30	6.53		0.000	10060	105725	2	
31	7.39	fluoride	59.797	1711949	25646393	2	0.00
32	8.11	glycolate	102.457	480784	9121965	2	0.00
33	8.61	acetate	122.149	609372	12208788	2	0.00
34	9.36	formate	129.387	1090237	21778151	2	0.00
35	11.31		0.000	7538495	271547782	2	
36	11.68		0.000	5436928	69966462	2	
37	12.00		0.000	6415299	512075938	3	
38	15.71		0.000	20689	1437928	4	
39	18.03		0.000	4994	121956	4	
40	18.61		0.000	5413	78334	4	

Totals 413.790 23502116 925586227

```

=====
Sample Name: BLANK                      Date: 12/23/1998 03:17:28
Data File  : C:\DX\DATA\98122201.D10
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10    Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----

```

```

External      1      1      5700 5Hz      0.00 19.00      500

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.13		0.000	14571	114680	1	
2	0.47		0.000	3073	47280	1	
3	1.30		0.000	7090	157420	1	
4	2.25		0.000	5009	32843	2	
5	2.47		0.000	2843	11177	2	
6	2.61		0.000	1880	8080	1	
7	2.80		0.000	5022	40253	2	
8	3.01		0.000	4653	34620	2	
9	3.17		0.000	4827	26333	2	
10	3.39		0.000	3289	26951	2	
11	3.59		0.000	4137	23117	2	
12	3.72		0.000	3168	17978	2	
13	3.96		0.000	3424	18716	2	
14	4.08		0.000	2950	25450	2	
15	4.37		0.000	378	1200	1	
16	4.48		0.000	825	1687	2	
17	4.56		0.000	3855	33433	2	
18	4.76		0.000	2329	13543	2	
19	4.87		0.000	1369	7868	2	
20	4.99		0.000	1614	6743	2	
21	5.12		0.000	3863	18958	2	
22	5.25		0.000	30313	224026	3	
23	5.51		0.000	3834	45260	4	
24	5.69		0.000	6677	109125	2	
25	6.03		0.000	3401	35678	2	
26	6.28		0.000	3696	55723	2	
27	6.85		0.000	5096	81660	1	
28	7.44	fluoride	0.010	4264	76964	3	0.00
29	7.87		0.000	347	2583	4	
30	8.05	glycolate	-0.014	3041	50733	3	0.00
31	8.48		0.000	384	1472	4	
32	8.64	acetate	-0.014	3343	82199	2	0.00
33	9.31	formate	0.010	5332	124284	2	0.00
34	9.79		0.000	5988	114549	2	
35	11.31		0.000	7413799	264814547	2	
36	11.71		0.000	5348512	67821925	2	
37	12.00		0.000	6139035	496942702	3	
39	17.95		0.000	1674	91008	4	
Totals			-0.008	19058907	831342771		

File: 98122201.D10 Sample: BLANK

Sample Name: S98T003384 SAM Date: 12/23/1998 04:35:20
 Data File : C:\DX\DATA\98122201.D13
 Method : C:\DX\METHOD\AAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 13 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	5700	5Hz	0.00	19.00		500

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	9456	49540	1	
2	0.33		0.000	25629	360366	2	
3	0.73		0.000	8346	105647	2	
4	1.50		0.000	200	2280	1	
5	2.23		0.000	7200	76800	1	
7	2.71		0.000	3364	30313	2	
8	2.81		0.000	5515	90039	3	
9	3.28		0.000	387	1123	4	
10	3.39		0.000	2117	5400	1	
11	3.80		0.000	1364	15288	2	
12	3.96		0.000	4658	60457	2	
13	4.39		0.000	1642	16056	2	
14	4.55		0.000	3936	60083	2	
15	4.95		0.000	819	3611	2	
16	5.12		0.000	2749	8343	2	
17	5.25		0.000	24572	165250	2	
18	5.48		0.000	168	1120	1	
19	5.56		0.000	952	6566	2	
20	5.69		0.000	3933	62198	2	
21	6.15		0.000	728	6233	2	
22	6.29		0.000	3289	39102	2	
23	6.85		0.000	5474	39073	2	
24	7.36	fluoride	211.962	78301	1432081	2	0.00
25	8.61	acetate	1356.912	113604	2841977	2	0.00
26	9.33	formate	1050.582	152353	3378054	2	0.00
27	10.93		0.000	6404016	66249204	2	
28	11.20		0.000	34725294	416617463	3	
29	11.71		0.000	4708418	44923088	4	
30	11.95		0.000	61406070	1038617221	3	
31	17.23		0.000	25375	347067	4	
32	17.76		0.000	14911	569846	4	

Totals 2619.456 107744839 1576180891

```

=====
Sample Name: S98T003384 DUP                      Date: 12/23/1998 04:58:58
Data File  : C:\DX\DATA\98122201.D14
Method     : C:\DX\METHOD\AAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          5151   5700  5Hz   0.00 19.00      500
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.13		0.000	20111	231993	2	
2	0.73		0.000	4053	62719	2	
3	1.17		0.000	454	11130	2	
4	1.57		0.000	1290	6032	2	
5	2.20		0.000	400	840	1	
6	2.31		0.000	2600	14700	1	
7	2.51		0.000	2975	13782	2	
8	2.71		0.000	5300	61785	2	
9	2.91		0.000	4225	23975	2	
10	3.04		0.000	7375	44328	2	
11	3.17		0.000	7125	77528	2	
12	3.39		0.000	2525	11990	2	
13	3.76		0.000	3683	44880	3	
14	4.00		0.000	572	1655	4	
15	4.08		0.000	2546	13344	2	
16	4.21		0.000	5111	31917	2	
17	4.36		0.000	3913	17924	2	
18	4.45		0.000	3568	14266	2	
19	4.60		0.000	3970	19215	2	
20	4.68		0.000	6189	30865	2	
21	4.83		0.000	6791	136612	2	
22	5.25		0.000	39559	538703	3	
23	6.03		0.000	5452	105670	4	
24	6.77		0.000	7060	158587	2	
25	7.39	fluoride	315.150	120860	2300324	2	0.00
26	8.61	acetate	2365.230	185858	4785126	2	0.00
27	9.36	formate	1636.059	236466	5295044	2	0.00
28	10.91		0.000	8197725	79621159	2	
29	11.20		0.000	47686964	485875966	3	
30	11.68		0.000	4838595	45877089	4	
31	11.92		0.000	81567550	1275941280	2	
32	18.64		0.000	171	5000	1	
Totals			4316.438	142981037	1901375427		

Sample Name: S98T003384 SPK Date: 12/23/1998 05:21:02
 Data File : C:\DX\DATA\98122201.D15
 Method : C:\DX\METHOD\AAS14GLY.MBT
 ACI Address: 1 System: 2 Inject#: 15 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	5700	5Hz	0.00	19.00	500	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	40000	376160	1	
2	1.23		0.000	722	4720	1	
3	1.57		0.000	4343	62700	1	
4	2.23		0.000	3480	12083	2	
5	2.32		0.000	2529	15294	2	
6	2.71		0.000	6000	65360	2	
7	2.84		0.000	3067	9220	2	
8	2.99		0.000	2626	8002	2	
9	3.09		0.000	2948	19824	2	
10	3.17		0.000	3089	13305	2	
11	3.28		0.000	810	2307	2	
12	3.36		0.000	1814	12291	2	
13	3.60		0.000	1076	4820	2	
14	3.73		0.000	4710	64470	2	
15	4.16		0.000	2277	5591	2	
16	4.23		0.000	2754	8929	2	
17	4.32		0.000	1533	3960	1	
18	4.40		0.000	2587	26802	2	
19	4.71		0.000	2020	7901	2	
20	4.91		0.000	12155	127990	2	
21	5.25		0.000	44176	522306	3	
22	5.84		0.000	2464	42456	4	
23	6.05		0.000	7935	133338	4	
24	7.39	fluoride	3441.167	1942083	29044337	2	0.00
25	8.11	glycolate	5637.153	516397	9845602	2	0.00
26	8.61	acetate	8787.637	850617	17102679	2	0.00
27	9.36	formate	7945.012	1328943	26324552	2	0.00
28	10.91		0.000	8181616	87475127	2	
29	11.20		0.000	52328744	522771457	3	
30	11.68		0.000	4930033	46497505	4	
31	11.92		0.000	94429057	1441794415	3	
32	16.32		0.000	29075	2197660	4	
Totals			25810.969	164691680	2184603164		

File: 98122201.D15 Sample: S98T003384 SPK

LABCORE Completed Worklist Report for Worklist# 27636

Analyst: adp

Instrument: IC40S1

Book#: 29N21D

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 @IC-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	③IC-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	③IC-QC CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	③IC-QC NO2	QC	1	1.60e-01	0.160	ug/mL
1 CCB	0	③IC-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	③IC-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	③IC-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	③IC-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	③IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	③IC-QC F	QC	5.30e1	6.53e+01	110.678 % Recovery	
2 LCS-INST	0	③IC-QC CL	QC	8.00e1	8.04e+01	100.500 % Recovery	
2 LCS-INST	0	③IC-QC NO2	QC	5.38e2	5.29e+02	98.327 % Recovery	
2 LCS-INST	0	③IC-QC BR	QC	5.86e2	5.64e+02	96.246 % Recovery	
2 LCS-INST	0	③IC-QC NO3	QC	5.92e2	5.73e+02	96.791 % Recovery	
2 LCS-INST	0	③IC-QC P04	QC	5.47e2	5.14e+02	93.967 % Recovery	
2 LCS-INST	0	③IC-QC S04	QC	6.38e2	6.16e+02	96.552 % Recovery	
2 LCS-INST	0	③IC-QC OXALATE2	QC	5.30e2	5.32e+02	100.377 % Recovery	
3 CCV	0	③IC-QC F	QC	6.40e1	6.42e+01	100.312 % Recovery	
3 CCV	0	③IC-QC CL	QC	9.00e1	8.70e+01	96.667 % Recovery	
3 CCV	0	③IC-QC NO2	QC	5.34e2	5.14e+02	96.255 % Recovery	
3 CCV	0	③IC-QC BR	QC	6.30e2	5.95e+02	94.444 % Recovery	
3 CCV	0	③IC-QC NO3	QC	6.98e2	6.47e+02	92.693 % Recovery	
3 CCV	0	③IC-QC P04	QC	6.32e2	5.82e+02	92.089 % Recovery	
3 CCV	0	③IC-QC S04	QC	6.99e2	6.56e+02	93.848 % Recovery	
3 CCV	0	③IC-QC OXALATE2	QC	5.12e2	4.82e+02	94.141 % Recovery	
4 SAMPLE	S98T003384	③IC-01 F-02	LIQUID	N/A	3.058e+02	61.610 ug/mL	
4 SAMPLE	S98T003384	③IC-01 CL-02	LIQUID	N/A	4.364e+03	87.570 ug/mL	
4 SAMPLE	S98T003384	③IC-01 NO2-02	LIQUID	N/A	6.471e+04	556.300 ug/mL	
4 SAMPLE	S98T003384	③IC-01 BR-02	LIQUID	N/A	6.759e+02	643.900 ug/mL	
4 SAMPLE	S98T003384	③IC-01 NO3-02	LIQUID	N/A	1.823e+05	716.000 ug/mL	
4 SAMPLE	S98T003384	③IC-01 P04-02	LIQUID	N/A	1.050e+03	618.100 ug/mL	
4 SAMPLE	S98T003384	③IC-01 S04-02	LIQUID	N/A	2.551e+03	710.800 ug/mL	
4 SAMPLE	S98T003384	③IC-01 OXALATE2	LIQUID	N/A	< 5.409e+02	540.900 ug/mL	
5 DUP	S98T003384	③IC-01 F-02	LIQUID	3.06e+02	3.06e+02	0.000 RPD	
5 DUP	S98T003384	③IC-01 CL-02	LIQUID	4.36e+03	4.28e+03	1.852 RPD	
5 DUP	S98T003384	③IC-01 NO2-02	LIQUID	6.47e+04	6.29e+04	2.821 RPD	
5 DUP	S98T003384	③IC-01 BR-02	LIQUID	6.76e+02	6.95e+02	2.772 RPD	
5 DUP	S98T003384	③IC-01 NO3-02	LIQUID	1.82e+05	1.81e+05	0.551 RPD	
5 DUP	S98T003384	③IC-01 P04-02	LIQUID	1.05e+03	9.57e+02	9.268 RPD	
5 DUP	S98T003384	③IC-01 S04-02	LIQUID	2.55e+03	2.60e+03	1.942 RPD	
5 DUP	S98T003384	③IC-01 OXALATE2	LIQUID	<5.41e2	5.90e+02	RPD	
6 SPK	S98T003384	③IC-01 F-02	LIQUID	5.90e1	6.91e+01	117.119 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27636

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SPK	S98T003384	0	@IC-01	CL-02	LIQUID	8.00e1	8.00e+01	100.000 % Recovery
6 SPK	S98T003384	0	@IC-01	NO2-02	LIQUID	5.38e2	5.44e+02	101.115 % Recovery
6 SPK	S98T003384	0	@IC-01	BR-02	LIQUID	5.86e2	5.10e+02	87.031 % Recovery
6 SPK	S98T003384	0	@IC-01	NO3-02	LIQUID	5.92e2	5.15e+02	86.993 % Recovery
6 SPK	S98T003384	0	@IC-01	PO4-02	LIQUID	5.47e2	4.84e+02	88.493 % Recovery
6 SPK	S98T003384	0	@IC-01	SO4-02	LIQUID	6.38e2	5.96e+02	93.417 % Recovery
6 SPK	S98T003384	0	@IC-01	OXALATE2	LIQUID	5.30e2	5.18e+02	97.736 % Recovery

Final page for worklist# 27636

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 12/28/98

LABCORE Data Entry Template for Worklist# 27636

Analyst: ADP Instrument: IC 4051 Book# 29N21-D LCS
Method: LA-533-105 Rev/Mod FC 28N21-D CCV
Worklist Comment: BY-105 @IC-01 RLC

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	SAMPLE	S98T003384 0			@IC-01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
5	DUP	S98T003384 0			@IC-01	LIQUID		
6	SPK	S98T003384 0			@IC-01	LIQUID		

Final page for worklist # 27636

Anthony Perinetti 12-19-98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

up loaded 12-21-98
John Howell

Validated 12/28/98 JMF-ey

27636DEC.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 29N21-D LCS

Date: 12/19/1998 09:16:11

Data File : C:\DX\DATA\98121901.D03

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 3

Detector: CDM-1

Analyst: *Anthony Purvis* Column: AG4A/AS4A anion column

12-19-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	82	346	2	
2	1.01	fluoride	65.306/10.69	2193	10910	2	1.33
3	1.48	chloride	80.380/100.48	2442	10478	1	0.68
4	1.75	nitrite	98.29 528.821/89.33	9251	46654	1	-0.19
5	2.55	bromide	564.198/96.28	6261	35547	1	-1.42
6	2.88	nitrate	572.832/96.76	7407	49369	1	-1.71
7	4.18	phosphate	513.958/93.96	1585	19147	1	1.70
8	5.39	sulfate	96.53 615.841/42.84	4808	63334	1	1.06
9	6.99	oxalate	531.814/100.35	2483	41733	1	0.82
Totals			3473.150	36513	277516		

File: 98121901.D03 Sample: 29N21-D LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1395 TO 1400

Sample Name: 28N21-D CCV Date: 12/19/1998 09:40:17
 Data File : C:\DX\DATA\98121901.D04
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 4 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	77	279	1	
2	1.00	fluoride	64.211	2129	10722	1	0.00
3	1.45	chloride	87.024	2608	11358	1	-1.13
4	1.72	nitrite	514.291	9231	45345	1	-1.71
5	2.51	bromide	594.578	6670	37522	1	-2.96
6	2.83	nitrate	646.554	8487	55954	1	-3.30
7	4.13	phosphate	581.869	1821	21765	1	0.41
8	5.33	sulfate	655.561	5528	67552	1	0.06
9	6.93	oxalate	481.872	2398	37767	1	0.05
Totals			3625.960	38949	288266		

File: 98121901.D04 Sample: 28N21-D CCV

Sample Name: BLANK

Date: 12/19/1998 09:00:01

Data File : C:\DX\DATA\98121901.D02

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 2

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.68	nitrite	0.160	57	635	1	-4.00
2	3.43		0.000	28	154	1	
3	6.24	oxalate	0.073	23	378	1	-9.96
Totals			0.233	108	1167		

File: 98121901.D02 Sample: BLANK

Sample Name: S98T003384 SAM Date: 12/19/1998 10:12:57
Data File : F:\DATA\98121901.D06
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 6 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject
External 1 5151 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	380	2005	2	
2	1.01	fluoride	305.784	231	892	2	1.33
3	1.09		0.000	419	3167	2	
4	1.44	chloride	4363.565	2362	11164	1	-2.04
5	1.71	nitrite	64712.522	21229	113667	1	-2.10
6	2.51	bromide	675.888	48	225	1	-3.22
7	2.75	nitrate	182291.178	42846	337252	1	-0.12
8	3.69		0.000	32	257	1	
9	4.13	phosphate	1050.147	43	453	1	0.57
10	5.33	sulfate	2551.000	292	4162	1	0.06
11	6.99	oxalate	359.152	25	353	1	0.82
Totals			256309.236	67907	473599		

File: 98121901.D06 Sample: S98T003384 SAM

Data Reprocessed On 12/21/1998 08:42:01

```

=====
Sample Name: S98T003384 DUP                      Date: 12/19/1998 10:24:38
Data File  : F:\DATA\98121901.D07
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 7              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1         5151   3000  5Hz   0.00 10.00         30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	395	2068	2	
2	1.02	fluoride	306.171	235	893	2	2.00
3	1.09		0.000	416	3020	2	
4	1.44	chloride	4283.665	2385	10956	1	-2.04
5	1.71	nitrite	62886.029	21444	110408	1	-2.10
6	2.51	bromide	695.476	52	250	1	-3.22
7	2.75	nitrate	181130.462	42987	334893	1	-0.12
8	4.13	phosphate	957.349	40	385	1	0.57
9	5.33	sulfate	2603.777	295	4269	1	0.06
10	6.99	oxalate	589.630	21	707	1	0.82
Totals			253452.559	68271	467849		

File: 98121901.D07 Sample: S98T003384 DUP

Data Reprocessed On 12/21/1998 08:42:53

Sample Name: S98T003384 SPK	Date: 12/19/1998 10:36:00
Data File : F:\DATA\98121901.D08	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 8	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	411	1936	2	
2	1.00	fluoride	3796.419	1853	12481	2	0.00
3	1.43	chloride	8402.738	4415	21784	1	-2.49
4	1.72	nitrite	92211.854	30037	162940	1	-1.71
5	2.49	bromide	26422.632	6304	32554	1	-3.73
6	2.73	nitrate	208300.692	49111	390975	1	0.12
7	3.69		0.000	59	751	1	
8	4.11	phosphate	25485.178	1548	18599	1	0.08
9	5.33	sulfate	32645.162	5587	65909	1	0.06
10	6.88	oxalate	26181.437	2302	40267	1	-0.72
Totals			423446.112	101626	748197		

File: 98121901.D08 Sample: S98T003384 SPK

LABCORE Completed Worklist Report for Worklist# 27686

Analyst: adp

Instrument: IC40S1

Book#: 51N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	6.80e-02	0.068	ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC SO4	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	LCS-INST	0	@IC-QC F	QC	5.32e1	6.05e+01	102.196 % Recovery	
2	LCS-INST	0	@IC-QC CL	QC	7.96e1	7.78e+01	97.739 % Recovery	
2	LCS-INST	0	@IC-QC NO2	QC	5.35e2	4.56e+02	85.234 % Recovery	
2	LCS-INST	0	@IC-QC BR	QC	5.86e2	5.41e+02	92.321 % Recovery	
2	LCS-INST	0	@IC-QC NO3	QC	5.92e2	5.93e+02	100.169 % Recovery	
2	LCS-INST	0	@IC-QC P04	QC	5.48e2	5.20e+02	94.891 % Recovery	
2	LCS-INST	0	@IC-QC SO4	QC	6.38e2	5.98e+02	93.730 % Recovery	
2	LCS-INST	0	@IC-QC OXALATE2	QC	5.52e2	5.65e+02	102.355 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.54e+01	102.188 % Recovery	
3	CCV	0	@IC-QC CL	QC	8.95e1	8.90e+01	99.441 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.37e2	5.30e+02	98.696 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.37e+02	101.111 % Recovery	
3	CCV	0	@IC-QC NO3	QC	7.00e2	7.07e+02	101.000 % Recovery	
3	CCV	0	@IC-QC P04	QC	6.35e2	6.50e+02	102.362 % Recovery	
3	CCV	0	@IC-QC SO4	QC	7.00e2	6.84e+02	97.714 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.33e2	5.66e+02	106.191 % Recovery	
4	BLNK-PREP	0	@IC-01 F-02	SOLID	1	<1.20e-2		ug/g
4	BLNK-PREP	0	@IC-01 CL-02	SOLID	1	<1.70e-2		ug/g
4	BLNK-PREP	0	@IC-01 NO2-02	SOLID	1	1.21e-01	0.121	ug/g
4	BLNK-PREP	0	@IC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4	BLNK-PREP	0	@IC-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4	BLNK-PREP	0	@IC-01 P04-02	SOLID	1	<1.20e-1		ug/g
4	BLNK-PREP	0	@IC-01 SO4-02	SOLID	1	<1.38e-1		ug/g
4	BLNK-PREP	0	@IC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5	SAMPLE	S98T002714	0 W @IC-01 NO2-02	SOLID	N/A	1.999e+04	868.400	ug/g
5	SAMPLE	S98T002714	0 W @IC-01 BR-02	SOLID	N/A	1.005e+03	1005.000	ug/g
5	SAMPLE	S98T002714	0 W @IC-01 SO4-02	SOLID	N/A	2.603e+04	1109.000	ug/g
5	SAMPLE	S98T002714	0 W @IC-01 OXALATE2	SOLID	N/A	1.352e+04	844.200	ug/g
6	DUP	S98T002714	0 W @IC-01 F-02	SOLID	1.24e+04	7.43e+03	50.126	RPD
6	DUP	S98T002714	0 W @IC-01 CL-02	SOLID	1.52e+03	1.23e+03	21.091	RPD
6	DUP	S98T002714	0 W @IC-01 NO2-02	SOLID	2.00e+04	1.84e+04	8.333	RPD
6	DUP	S98T002714	0 W @IC-01 BR-02	SOLID	<1.00e3	<9.92e2		RPD
6	DUP	S98T002714	0 W @IC-01 NO3-02	SOLID	6.72e+04	5.32e+04	23.256	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27686

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002714	0 W	@IC-01	PO4-02	SOLID	6.59e+04	1.45e+04	127.861 RPD
6 DUP	S98T002714	0 W	@IC-01	SO4-02	SOLID	2.60e+04	2.93e+04	11.935 RPD
6 DUP	S98T002714	0 W	@IC-01	OxALATE2	SOLID	1.35e+04	1.35e+04	0.000 RPD
7 SAMPLE	S98T002741	0 W	@IC-01	F-02	SOLID	N/A	2.649e+03	122.600 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	CL-02	SOLID	N/A	5.770e+02	173.700 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	NO2-02	SOLID	N/A	7.487e+03	1103.000 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	BR-02	SOLID	N/A	1.277e+03	1277.000 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	NO3-02	SOLID	N/A	3.627e+05	1420.000 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	PO4-02	SOLID	N/A	2.011e+04	1226.000 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	SO4-02	SOLID	N/A	4.664e+03	1410.000 ug/g
7 SAMPLE	S98T002741	0 W	@IC-01	OxALATE2	SOLID	N/A	8.560e+03	1073.000 ug/g
8 DUP	S98T002741	0 W	@IC-01	F-02	SOLID	2.65e+03	2.43e+03	8.661 RPD
8 DUP	S98T002741	0 W	@IC-01	CL-02	SOLID	5.77e+02	6.26e+02	8.146 RPD
8 DUP	S98T002741	0 W	@IC-01	NO2-02	SOLID	7.49e+03	8.00e+03	6.585 RPD
8 DUP	S98T002741	0 W	@IC-01	BR-02	SOLID	<1.28e3	<1.29e3	RPD
8 DUP	S98T002741	0 W	@IC-01	NO3-02	SOLID	3.63e+05	3.39e+05	6.838 RPD
8 DUP	S98T002741	0 W	@IC-01	PO4-02	SOLID	2.01e+04	1.63e+04	20.879 RPD
8 DUP	S98T002741	0 W	@IC-01	SO4-02	SOLID	4.66e+03	5.16e+03	10.183 RPD
8 DUP	S98T002741	0 W	@IC-01	OxALATE2	SOLID	8.56e+03	9.18e+03	6.990 RPD
9 SPK	S98T002741	0 W	@IC-01	F-02	SOLID	5.92e1	6.50e+01	109.797 % Recovery
9 SPK	S98T002741	0 W	@IC-01	CL-02	SOLID	7.96e1	7.65e+01	96.106 % Recovery
9 SPK	S98T002741	0 W	@IC-01	NO2-02	SOLID	5.35e2	4.71e+02	88.037 % Recovery
9 SPK	S98T002741	0 W	@IC-01	BR-02	SOLID	5.86e2	5.51e+02	94.027 % Recovery
9 SPK	S98T002741	0 W	@IC-01	NO3-02	SOLID	5.92e2	5.91e+02	99.831 % Recovery
9 SPK	S98T002741	0 W	@IC-01	PO4-02	SOLID	5.48e2	5.51e+02	100.547 % Recovery
9 SPK	S98T002741	0 W	@IC-01	SO4-02	SOLID	6.38e2	6.14e+02	96.238 % Recovery
9 SPK	S98T002741	0 W	@IC-01	OxALATE2	SOLID	5.52e2	5.83e+02	105.616 % Recovery

Final page for worklist# 27686

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

[Signature]
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27686

Analyst: ADP Instrument: IC 4051 Book# 5021C LCS
5021C CCV

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002714 0 W		@IC-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
6 DUP	S98T002714 0 W		@IC-01	SOLID		
7 SAMPLE	S98T002741 0 W		@IC-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
8 DUP	S98T002741 0 W		@IC-01	SOLID		
9 SPK	S98T002741 0 W		@IC-01	SOLID		

Final page for worklist # 27686

See attached
 Signature Date

Signature Date

Data Entry Comments:

uploaded 2-17-99
John W. Wall
27686 FEB. CSV
2/18/99 gmd
Same analytes rejected from
S98T002714 test validated
Reanalyzed S98T002714 on W# 28341. f. 3/24/99
1403

12/28/98 14:06
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27686

Analyst: ADP Instrument: IC 4051 Book# 51N21-C LCS
 Method: LA-533-105 Rev/Mod F-0 50N21-C CCV

Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002714 0 W	@IC-01	SOLID	98000436 BY-105 (2)		
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
6 DUP	S98T002714 0 W	@IC-01	SOLID			
7 SAMPLE	S98T002741 0 W	@IC-01	SOLID	98000490 BY-105 (2)		
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
8 DUP	S98T002741 0 W	@IC-01	SOLID			
9 SPK	S98T002741 0 W	@IC-01	SOLID			
10 SAMPLE	S98T002848 0 W	@IC-01	SOLID	98000437 BY-105 (2)		
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
11 DUP	S98T002848 0 W	@IC-01	SOLID			
12 SPK	S98T002848 0 W	@IC-01	SOLID			

Final page for worklist # 27686

Anthony Parvuta 2-7-99
 Analyst Signature Date

Put these on new wkl miss identified
or run sheets
 Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

=====

Sample Name: 51N21-C LCS	Date: 02/07/1999 00:48:51
Data File : C:\DX\DATA\99020601.D19	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: A System: 1 Inject#: 19	Detector: CDM-1
Analyst: <i>Anthony Purts</i>	Column: AG4A/AS4A anion column

=====

6-3-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	72	261	2	
2	0.99	fluoride	60.483	102.17	1875	9181	2 -0.34
3	1.42	chloride	77.806	97.75	2153	9575	1 -0.70
4	1.67	nitrite	455.955	85.23	7768	38596	1 -2.14
5	2.43	bromide	541.144	92.35	5767	31216	1 -4.95
6	2.74	nitrate	593.249	100.21	7375	46415	1 -0.72
7	4.01	phosphate	520.164	94.92	1433	17171	1 1.35
8	5.17	sulfate	597.809	93.70	4983	57812	1 3.26
9	6.67	oxalate	564.802	102.32	2256	38500	1 3.31
Totals			3411.411	33681	248728		

File: 99020601.D19 Sample: 51N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1405 TO 1413.

```

=====
Sample Name: 50N21-C CCV                      Date: 02/07/1999 01:00:51
Data File  : C:\DX\DATA\99020601.D20
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 20           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          101    3000  5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	84	316	2	
2	0.99	fluoride	65.414	1915	9942	2	-0.34
3	1.43	chloride	89.000	2590	10973	1	-0.23
4	1.68	nitrite	530.441	9132	44920	1	-1.75
5	2.44	bromide	637.054	6663	36848	1	-4.69
6	2.74	nitrate	706.580	8406	55431	1	-0.72
7	4.03	phosphate	650.109	1795	21563	1	1.68
8	5.17	sulfate	684.335	5416	66318	1	3.26
9	6.72	oxalate	566.298	2505	38603	1	4.13
Totals			3929.232	38506	284913		

File: 99020601.D20 Sample: 50N21-C CCV

Sample Name: BLANK Date: 02/06/1999 23:51:32
Data File : C:\DX\DATA\99020601.D18
Method : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 18 Detector: CDM-1
Analyst : Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	34	150	1	
2	1.36	chloride	0.068	92	833	1	-4.90
3	2.79	nitrate	0.009	13	66	1	0.97
Totals			0.076	140	1050		

File: 99020601.D18 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 02/07/1999 01:12:51
Data File  : C:\DX\DATA\99020601.D21
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 21         Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.27		0.000	56	518	1	
2	1.44	chloride	0.013	41	162	1	0.70
3	1.69	nitrite	0.121	231	1029	1	-1.36
4	2.79	nitrate	0.036	52	281	1	1.21
Totals			0.170	380	1990		

File: 99020601.D21 Sample: PREP BLANK

Data Reprocessed On 02/08/1999 13:55:31

```

=====
Sample Name: S98T002714 SAM                      Date: 02/07/1999 01:25:56
Data File  : F:\DATA\99020601.D22
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 22              Detector:CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	120	454	2	
2	0.99	fluoride	63.188	4863	24214	2	-0.34
3	1.42	chloride	7.734	521	2322	1	-0.70
4	1.67	nitrite	101.944	4194	21233	1	-2.14
5	2.73	nitrate	342.383	10196	66384	1	0.12
6	4.02	phosphate	335.939	2293	27630	1	1.52
7	5.17	sulfate	132.728	2404	31416	1	3.26
8	6.72	oxalate	68.963	713	11518	1	4.13
Totals			1052.880	25304	185171		

File: 99020601.D22 Sample: S98T002714 SAM

```

=====
Sample Name: S98T002714  DUP                      Date: 02/07/1999 01:38:19
Data File  : C:\DX\DATA\99020601.D23
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 23                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	154	555	2	
2	0.99	fluoride	38.377	2924	14474	2	-0.34
3	1.42	chloride	6.370	431	1912	1	-0.70
4	1.68	nitrite	94.998	3894	19784	1	-1.75
5	2.74	nitrate	274.819	8051	53072	1	-0.72
6	4.04	phosphate	74.946	514	6022	1	2.02
7	5.17	sulfate	151.389	2656	35872	1	3.26
8	6.72	oxalate	69.509	703	11610	1	4.13

Totals 710.408 19326 143301

File: 99020601.D23 Sample: S98T002714 DUP

Data Reprocessed On 02/08/1999 13:56:19

```

=====
Sample Name: S98T002741 SAM                      Date: 02/07/1999 02:03:56
Data File  : F:\DATA\99020601.D25
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 25              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51    3000   5Hz   0.00 10.00         30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	113	441	2	
2	0.99	fluoride	13.228	801	3943	2	-0.34
3	1.42	chloride	2.881	162	694	1	-0.70
4	1.67	nitrite	37.381	1239	6253	1	-2.14
5	2.67	nitrate	1811.117	40358	303968	1	0.12
6	4.03	phosphate	100.405	537	6489	1	1.68
7	5.17	sulfate	23.285	311	4402	1	3.26
8	6.72	oxalate	42.739	359	5732	1	4.13
Totals			2031.037	43880	331922		

File: 99020601.D25 Sample: S98T002741 SAM

Data Reprocessed On 02/08/1999 13:57:07

```

=====
Sample Name: S98T002741  DUP                      Date: 02/07/1999 02:15:00
Data File  : F:\DATA\99020601.D26
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 26                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51    3000   5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	114	444	2	
2	0.99	fluoride	12.037	731	3586	2	-0.34
3	1.42	chloride	3.097	173	746	1	-0.70
4	1.67	nitrite	39.624	1326	6628	1	-2.14
5	2.67	nitrate	1676.138	37381	279069	1	0.12
6	4.04	phosphate	80.558	436	5199	1	2.02
7	5.17	sulfate	25.564	346	4833	1	3.26
8	6.72	oxalate	45.472	376	6099	1	4.13
Totals			1882.491	40882	306604		

File: 99020601.D26 Sample: S98T002741 DUP

Data Reprocessed On 02/08/1999 13:57:49

```

=====
Sample Name: S98T002741 SPK                      Date: 02/07/1999 02:26:01
Data File  : F:\DATA\99020601.D27
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 27              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	173	627	2	
2	0.99	fluoride	45.751	2754	13858	2	-0.34
3	1.42	chloride	41.134	2239	10031	1	-0.70
4	1.68	nitrite	272.914	9185	45773	1	-1.75
5	2.43	bromide	275.320	6060	31457	1	-0.14
6	2.66	nitrate	2106.536	46798	360076	1	0.00
7	4.03	phosphate	376.137	2055	24794	1	1.68
8	5.17	sulfate	330.047	5241	63295	1	3.26
9	6.67	oxalate	334.245	2608	45182	1	3.31
Totals			3782.084	77113	595091		

File: 99020601.D27 Sample: S98T002741 SPK

LABCORE Completed Worklist Report for Worklist# 27687

Analyst: kjt

Instrument: IC40S1

Book#: 50N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0 @IC-QC F QC			1	<1.20e-2		ug/mL
1 CCB	0 @IC-QC CL QC			1	<1.70e-2		ug/mL
1 CCB	0 @IC-QC NO2 QC			1	<1.08e-1		ug/mL
1 CCB	0 @IC-QC BR QC			1	<1.25e-1		ug/mL
1 CCB	0 @IC-QC NO3 QC			1	<1.39e-1		ug/mL
1 CCB	0 @IC-QC P04 QC			1	<1.20e-1		ug/mL
1 CCB	0 @IC-QC S04 QC			1	<1.38e-1		ug/mL
1 CCB	0 @IC-QC OXALATE2 QC			1	<1.05e-1		ug/mL
2 LCS-INST	0 @IC-QC F QC			5.92e1	5.95e+01	100.507 % Recovery	
2 LCS-INST	0 @IC-QC CL QC			7.96e1	7.79e+01	97.864 % Recovery	
2 LCS-INST	0 @IC-QC NO2 QC			5.15e2	4.86e+02	94.369 % Recovery	
2 LCS-INST	0 @IC-QC BR QC			5.86e2	5.58e+02	95.222 % Recovery	
2 LCS-INST	0 @IC-QC NO3 QC			5.92e2	5.68e+02	95.946 % Recovery	
2 LCS-INST	0 @IC-QC P04 QC			5.48e2	5.42e+02	98.905 % Recovery	
2 LCS-INST	0 @IC-QC S04 QC			6.38e2	6.15e+02	96.395 % Recovery	
2 LCS-INST	0 @IC-QC OXALATE2 QC			5.22e2	5.45e+02	104.406 % Recovery	
3 CCV	0 @IC-QC F QC			6.40e1	6.48e+01	101.250 % Recovery	
3 CCV	0 @IC-QC CL QC			8.95e1	8.97e+01	100.223 % Recovery	
3 CCV	0 @IC-QC NO2 QC			5.45e2	5.63e+02	103.303 % Recovery	
3 CCV	0 @IC-QC BR QC			6.30e2	6.29e+02	99.841 % Recovery	
3 CCV	0 @IC-QC NO3 QC			7.00e2	7.03e+02	100.429 % Recovery	
3 CCV	0 @IC-QC P04 QC			6.35e2	6.57e+02	103.465 % Recovery	
3 CCV	0 @IC-QC S04 QC			7.00e2	6.97e+02	99.571 % Recovery	
3 CCV	0 @IC-QC OXALATE2 QC			5.27e2	5.70e+02	108.159 % Recovery	
4 BLNK-PREP	0 @IC-01 F-02 SOLID			1	<1.20e-2		ug/g
4 BLNK-PREP	0 @IC-01 CL-02 SOLID			1	<1.70e-2		ug/g
4 BLNK-PREP	0 @IC-01 NO2-02 SOLID			1	1.24e-01	0.124	ug/g
4 BLNK-PREP	0 @IC-01 BR-02 SOLID			1	<1.25e-1		ug/g
4 BLNK-PREP	0 @IC-01 NO3-02 SOLID			1	<1.39e-1		ug/g
4 BLNK-PREP	0 @IC-01 P04-02 SOLID			1	<1.20e-1		ug/g
4 BLNK-PREP	0 @IC-01 S04-02 SOLID			1	<1.38e-1		ug/g
4 BLNK-PREP	0 @IC-01 OXALATE2 SOLID			1	<1.05e-1		ug/g
5 SAMPLE	S98T003390 0 W @IC-01 F-02 SOLID			N/A	1.441e+03	240.700	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 CL-02 SOLID			N/A	4.214e+02	341.000	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 NO2-02 SOLID			N/A	5.509e+03	2165.000	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 BR-02 SOLID			N/A	< 2.507e+03	2507.000	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 NO3-02 SOLID			N/A	6.386e+05	2788.000	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 S04-02 SOLID			N/A	5.428e+03	2767.000	ug/g
5 SAMPLE	S98T003390 0 W @IC-01 OXALATE2 SOLID			N/A	3.130e+03	2106.000	ug/g
6 DUP	S98T003390 0 W @IC-01 F-02 SOLID			1.44e+03	1.52e+03	5.405	RPD
6 DUP	S98T003390 0 W @IC-01 CL-02 SOLID			4.21e+02	4.53e+02	7.323	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27687

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T003390	0 W	QIC-01	NO2-02	SOLID	5.51e+03	5.49e+03	0.364 RPD
6 DUP	S98T003390	0 W	QIC-01	BR-02	SOLID	<2.51e3	<2.46e3	RPD
6 DUP	S98T003390	0 W	QIC-01	NO3-02	SOLID	6.39e+05	6.44e+05	0.779 RPD
6 DUP	S98T003390	0 W	QIC-01	PO4-02	SOLID	3.44e+03	4.81e+03	33.212 RPD
6 DUP	S98T003390	0 W	QIC-01	SO4-02	SOLID	5.43e+03	6.12e+03	11.948 RPD
6 DUP	S98T003390	0 W	QIC-01	OXALATE2	SOLID	3.13e+03	2.99e+03	4.575 RPD
7 SPK	S98T003390	0 W	QIC-01	F-02	SOLID	5.92e1	5.73e+01	96.791 % Recovery
7 SPK	S98T003390	0 W	QIC-01	CL-02	SOLID	7.96e1	6.97e+01	87.563 % Recovery
7 SPK	S98T003390	0 W	QIC-01	NO2-02	SOLID	5.15e2	4.41e+02	85.631 % Recovery
7 SPK	S98T003390	0 W	QIC-01	BR-02	SOLID	5.86e2	4.98e+02	84.983 % Recovery
7 SPK	S98T003390	0 W	QIC-01	NO3-02	SOLID	5.92e2	5.19e+02	87.669 % Recovery
7 SPK	S98T003390	0 W	QIC-01	PO4-02	SOLID	5.48e2	4.94e+02	90.146 % Recovery
7 SPK	S98T003390	0 W	QIC-01	SO4-02	SOLID	6.38e2	5.56e+02	87.147 % Recovery
7 SPK	S98T003390	0 W	QIC-01	OXALATE2	SOLID	5.22e2	4.88e+02	93.487 % Recovery
8 SAMPLE	S98T003395	0 W	QIC-01	F-02	SOLID	N/A	4.104e+02	245.200 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	CL-02	SOLID	N/A	4.019e+02	347.400 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	NO2-02	SOLID	N/A	3.169e+03	2207.000 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	BR-02	SOLID	N/A	< 2.555e+03	2554.000 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	NO3-02	SOLID	N/A	6.554e+05	2841.000 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	PO4-02	SOLID	N/A	< 2.452e+03	2452.000 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	SO4-02	SOLID	N/A	< 2.820e+03	2820.000 ug/g
8 SAMPLE	S98T003395	0 W	QIC-01	OXALATE2	SOLID	N/A	2.164e+03	2146.000 ug/g
9 DUP	S98T003395	0 W	QIC-01	F-02	SOLID	4.10e+02	3.72e+02	9.719 RPD
9 DUP	S98T003395	0 W	QIC-01	CL-02	SOLID	4.02e+02	<3.33e2	RPD
9 DUP	S98T003395	0 W	QIC-01	NO2-02	SOLID	3.17e+03	3.13e+03	1.270 RPD
9 DUP	S98T003395	0 W	QIC-01	BR-02	SOLID	<2.56e3	<2.45e3	RPD
9 DUP	S98T003395	0 W	QIC-01	NO3-02	SOLID	6.55e+05	6.72e+05	2.562 RPD
9 DUP	S98T003395	0 W	QIC-01	PO4-02	SOLID	<2.45e3	<2.35e3	RPD
9 DUP	S98T003395	0 W	QIC-01	SO4-02	SOLID	<2.82e3	<2.70e3	RPD
9 DUP	S98T003395	0 W	QIC-01	OXALATE2	SOLID	2.16e+03	<2.06e3	RPD
10 SPK	S98T003395	0 W	QIC-01	F-02	SOLID	5.92e1	5.72e+01	96.622 % Recovery
10 SPK	S98T003395	0 W	QIC-01	CL-02	SOLID	7.96e1	7.44e+01	93.467 % Recovery
10 SPK	S98T003395	0 W	QIC-01	NO2-02	SOLID	5.15e2	4.62e+02	89.709 % Recovery
10 SPK	S98T003395	0 W	QIC-01	BR-02	SOLID	5.86e2	5.23e+02	89.249 % Recovery
10 SPK	S98T003395	0 W	QIC-01	NO3-02	SOLID	5.92e2	4.18e+02	70.608 % Recovery
10 SPK	S98T003395	0 W	QIC-01	PO4-02	SOLID	5.48e2	4.86e+02	88.686 % Recovery
10 SPK	S98T003395	0 W	QIC-01	SO4-02	SOLID	6.38e2	6.00e+02	94.044 % Recovery
10 SPK	S98T003395	0 W	QIC-01	OXALATE2	SOLID	5.22e2	5.07e+02	97.126 % Recovery
11 SAMPLE	S98T003400	0 W	QIC-01	F-02	SOLID	N/A	5.189e+01	49.730 ug/g
11 SAMPLE	S98T003400	0 W	QIC-01	CL-02	SOLID	N/A	1.805e+03	70.430 ug/g
11 SAMPLE	S98T003400	0 W	QIC-01	NO2-02	SOLID	N/A	3.254e+04	447.500 ug/g
11 SAMPLE	S98T003400	0 W	QIC-01	BR-02	SOLID	N/A	< 5.180e+02	517.900 ug/g
11 SAMPLE	S98T003400	0 W	QIC-01	PO4-02	SOLID	N/A	1.257e+04	497.300 ug/g
12 DUP	S98T003400	0 W	QIC-01	F-02	SOLID	5.19e+01	<4.97e1	RPD
12 DUP	S98T003400	0 W	QIC-01	CL-02	SOLID	1.80e+03	1.51e+03	17.523 RPD
12 DUP	S98T003400	0 W	QIC-01	NO2-02	SOLID	3.25e+04	2.73e+04	17.391 RPD
12 DUP	S98T003400	0 W	QIC-01	BR-02	SOLID	<5.18e2	<5.18e2	RPD
12 DUP	S98T003400	0 W	QIC-01	NO3-02	SOLID	6.44e+04	5.16e+04	22.069 RPD
12 DUP	S98T003400	0 W	QIC-01	PO4-02	SOLID	1.26e+04	1.10e+04	13.559 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27687

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
12 DUP	S98T003400	0 W @IC-01	S04-02	SOLID	4.07e+03	2.56e+04	145.130	RPD
12 DUP	S98T003400	0 W @IC-01	OXALATE2	SOLID	2.02e+04	5.19e+04	87.933	RPD
13 SPK	S98T003400	0 W @IC-01	F-02	SOLID	6.40e1	6.62e+01	103.438	% Recovery
13 SPK	S98T003400	0 W @IC-01	CL-02	SOLID	8.95e1	7.72e+01	86.257	% Recovery
13 SPK	S98T003400	0 W @IC-01	NO2-02	SOLID	5.45e2	5.24e+02	96.147	% Recovery
13 SPK	S98T003400	0 W @IC-01	BR-02	SOLID	6.30e2	5.40e+02	85.714	% Recovery
13 SPK	S98T003400	0 W @IC-01	NO3-02	SOLID	7.00e2	6.25e+02	89.286	% Recovery
13 SPK	S98T003400	0 W @IC-01	P04-02	SOLID	6.35e2	5.50e+02	86.614	% Recovery
13 SPK	S98T003400	0 W @IC-01	S04-02	SOLID	7.00e2	6.15e+02	87.857	% Recovery
13 SPK	S98T003400	0 W @IC-01	OXALATE2	SOLID	5.27e2	5.59e+02	106.072	% Recovery

Final page for worklist# 27687

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jan M. Lyle
Reviewer Signature _____ Date *1/26/99*

*S98T003390 & S98T003400 reanalyzed on Wt# 280666,
#3390 for P04 & #3400 for NO₂, SO₄ & oxalate. *3/24/99**

12/28/98 14:12
A-0004-1

LABCORE Data Entry Template for Worklist# 27687

Analyst: KST Instrument: IC 4000 sys 1 Book# 51221B
 Method: LA-533-105 Rev/Mod E-O
 Worklist Comment: BY-105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T003390 O W			@IC-01	SOLID	98000490	BY-105 (2)
		Analytes Requested:	BR-02		CL-02		F-02	NO2-02 NO3-02
					OXALATE2, PO4-02		SO4-02	
6	DUP	S98T003390 O W			@IC-01	SOLID		
7	SPK	S98T003390 O W			@IC-01	SOLID		
8	SAMPLE	S98T003395 O W			@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested:	BR-02		CL-02		F-02	NO2-02 NO3-02
					OXALATE2, PO4-02		SO4-02	
9	DUP	S98T003395 O W			@IC-01	SOLID		
10	SPK	S98T003395 O W			@IC-01	SOLID		
11	SAMPLE	S98T003400 O W			@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested:	BR-02		CL-02		F-02	NO2-02 NO3-02
					OXALATE2, PO4-02		SO4-02	
12	DUP	S98T003400 O W			@IC-01	SOLID		
13	SPK	S98T003400 O W			@IC-01	SOLID		

Final page for worklist # 27687

Data Entry Comments:

uploaded 1-25-99SPK27687-JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

PO₄ rejected on 3390 &NO₃, SO₄ & Oxalate on 3400 rejected
due to high RPD, SPK on 3395NO₃ failed due to high level of NO₂ in

the detector seen. We will not rerun this one, but

will rerun the others.

Please have

Jann - check

S98T003395 - rename

before uploading.

couldn't get computer

to rename peak

- SAFETY -
"A Way of Life"

1417

1/25/99

12/28/98 14:12

A-0004-1

LABCORE Data Entry Template for Worklist# 27687

S Type	Sample#	R A Test	Matrix	Group#	Project	
Analyst Signature		Date		Analyst Signature		Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 01/22/1999 08:58:51
 Data File : C:\DX\DATA\99012201.D01
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 1 Detector: CDM-1
 Analyst : *KS Hornbake* Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	2.90	nitrate	0.030	41	230	1	0.35
Totals			0.030	41	230		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1419 TO 1431

=====

Sample Name: LCS 51N21B	Date: 01/22/1999 09:10:45
Data File : C:\DX\DATA\99012201.D02	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 2	Detector: CDM-1
Analyst : <i>KS Hornbake</i>	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

-----

External	1	101	3000	5Hz	0.00	10.00	30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	66	236	2	
2	1.01	fluoride	100.5020 59.531	1790	9034	2	1.68
3	1.46	chloride	97.8020 77.878	2194	9584	1	2.10
4	1.73	nitrite	94.4070 486.284	8249	41170	1	0.97
5	2.52	bromide	95.2020 557.685	5809	32186	1	-1.56
6	2.84	nitrate	96.0020 568.124	6786	44423	1	-1.73
7	4.10	phosphate	99.0020 542.445	1435	17921	1	3.54
8	5.23	sulfate	96.3020 614.671	5131	59467	1	4.32
9	6.77	oxalate	94.5020 545.447	2403	37171	1	4.96
Totals			3452.064	33862	251192		

File: 99012201.D02 Sample: LCS 51N21B

Sample Name: CCV 50N21B Date: 01/22/1999 09:23:09
 Data File : C:\DX\DATA\99012201.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3 Detector: CDM-1
 Analyst : *Ks Thorndike* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	74	271	2	
2	1.01	fluoride	101.3010 64.770	1895	9842	2	1.68
3	1.46	chloride	100.3010 89.687	2526	11059	1	2.10
4	1.73	nitrite	163.3010 562.863	9264	47675	1	0.97
5	2.51	bromide	99.4010 529.187	6522	36385	1	-1.82
6	2.83	nitrate	100.4010 702.857	8208	55134	1	-2.19
7	4.09	phosphate	109.3010 657.162	1754	21803	1	3.37
8	5.23	sulfate	99.5010 696.683	5902	67535	1	4.32
9	6.77	oxalate	106.3010 570.336	2541	38880	1	4.96
Totals			3973.545	38686	288583		

File: 99012201.D03 Sample: CCV 50N21B

```

=====
Sample Name: PREP-BLK                      Date: 01/22/1999 09:36:56
Data File  : C:\DX\DATA\99012201.D04.
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 4         Detector: CDM-1
Analyst    : KS Herndike Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.31		0.000	54	531	1	
2	1.73	nitrite	0.124	116	1060	1	0.97
3	2.87	nitrate	0.018	27	139	1	-0.81
Totals			0.142	197	1729		

File: 99012201.D04 Sample: PREP-BLK

Data Reprocessed On 01/22/1999 11:15:35

```

=====
Sample Name: S98T003390 SAM                      Date: 01/22/1999 11:13:45
Data File  : C:\DX\DATA\99012201.D06
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 6                Detector: CDM-1
Analyst    : K5 Thorndike Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101      3000 5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	80	373	2	
2	1.01	fluoride	7.256	209	1087	2	1.68
3	1.46	chloride	2.122	58	258	1	2.10
4	1.72	nitrite	27.745	461	2343	1	3.58
5	2.77	nitrate	3215.859	35286	269540	1	-0.12
6	4.11	phosphate	17.342	55	562	1	3.87
7	5.23	sulfate	27.333	191	2608	1	4.32
8	6.77	oxalate	15.761	66	1066	1	4.96
Totals			3313.419	36407	277838		

File: 99012201.D06 Sample: S98T003390 SAM

Data Reprocessed On 01/22/1999 11:26:30

```

=====
Sample Name: S98T003390 DUP                      Date: 01/22/1999 11:24:30
Data File  : C:\DX\DATA\99012201.D07
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 7                Detector: CDM-1
Analyst    : Ks Kordike Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000 5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	77	327	2	
2	1.01	fluoride	7.824	228	1172	2	1.68
3	1.45	chloride	2.323	69	282	1	1.63
4	1.72	nitrite	28.173	476	2379	1	0.58
5	2.76	nitrate	3307.793	36041	277997	1	0.00
6	4.10	phosphate	24.693	69	801	1	3.54
7	5.23	sulfate	31.437	230	3000	1	4.32
8	6.77	oxalate	15.336	68	1038	1	4.96
Totals			3417.579	37258	286996		

File: 99012201.D07 Sample: S98T003390 DUP

Data Reprocessed On 01/22/1999 11:38:29

=====

Sample Name: S98T003390 SPK	Date: 01/22/1999 11:35:56
Data File : C:\DX\DATA\99012201.D08	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 8	Detector: CDM-1
Analyst : <i>KS Thorndike</i>	Column: AG4A/AS4A anion column

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	134	554	2	
2	1.01	fluoride	64.575	1810	9812	2	1.68
3	1.46	chloride	71.834	1963	8831	1	2.10
4	1.73	nitrite	468.756	7850	39682	1	0.97
5	2.51	bromide	497.909	5411	28688	1	-2.08
6	2.75	nitrate	3734.934	40672	317966	1	0.12
7	4.09	phosphate	511.325	1354	16874	1	3.37
8	5.23	sulfate	583.811	4938	56439	1	4.32
9	6.77	oxalate	503.319	2269	34280	1	4.96
Totals			6436.463	66401	513127		

File: 99012201.D08 Sample: S98T003390 SPK

Data Reprocessed On 01/25/1999 10:26:08

```

=====
Sample Name: S98T003395 SAM                      Date: 01/22/1999 11:47:07
Data File  : F:\DATA\99012201.D09
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 9                Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	73	340	2	
2	0.99	fluoride	2.028	61	304	2	0.34
3	1.44	chloride	1.986	48	241	1	0.70
4	1.70	nitrite	15.659	269	1322	1	-0.58
5	2.73	nitrate	3238.893	35783	271654	1	-0.12
6	4.07	phosphate	4.183	19	136	1	2.69
7	5.23	sulfate	5.584	45	533	1	4.32
8	6.72	oxalate	10.694	38	724	1	4.13
Totals			3279.027	36335	275253		

File: 99012201.D09 Sample: S98T003395 SAM

Data Reprocessed On 01/22/1999 13:14:10

```

=====
Sample Name: S98T003395 DUP      Date: 01/22/1999 11:57:53
Data File  : C:\DX\DATA\99012201.D10
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 10      Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    2783  5Hz   0.00  9.27        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	85	399	2	
2	1.00	fluoride	1.919	63	287	2	1.01
3	1.45	chloride	1.083	35	132	1	1.63
4	1.71	nitrite	16.128	273	1362	1	0.19
5	2.75	nitrate	3464.685	37989	292546	1	-0.12
6	5.23	sulfate	6.782	53	647	1	4.32
7	6.83	oxalate	7.187	28	486	1	5.79
Totals			3497.784	38525	295858		

File: 99012201.D10 Sample: S98T003395 DUP

Data Reprocessed On 01/22/1999 14:04:26

```

=====
Sample Name: S98T003395 SPK                      Date: 01/22/1999 14:01:58
Data File  : C:\DX\DATA\99012201.D13
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 13              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	139	568	2	
2	1.01	fluoride	59.226	1580	8987	2	2.36
3	1.46	chloride	76.400	2029	9400	1	2.10
4	1.72	nitrite	478.002	7742	40467	1	0.58
5	2.51	bromide	522.611	5764	30132	1	-2.08
6	2.75	nitrate	3657.210	38452	310608	1	-0.12
7	4.09	phosphate	485.599	1308	16010	1	3.37
8	5.23	sulfate	599.601	5064	57988	1	4.32
9	6.77	oxalate	518.173	2260	35299	1	4.96
Totals			6396.821	64337	509459		

File: 99012201.D13 Sample: S98T003395 SPK

HNF-1663 REV. 0

Sample Name: S98T003400

Date: 01/22/1999 14:21:46

Data File : C:\DX\DATA\99012201.D15

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 15

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	180	867	1	
2	1.03	fluoride	0.263	68	190	1	4.38
3	1.45	chloride	9.149	1161	5385	1	1.63
4	1.72	nitrite	164.919	12743	67271	1	0.58
5	2.79	nitrate	326.427	17806	125810	1	-3.58
6	4.09	phosphate	63.688	786	10035	1	3.37
7	5.23	sulfate	20.640	771	9487	1	4.32
8	6.72	oxalate	102.124	1919	33447	1	4.13
Totals			687.210	35434	252491		

File: 99012201.D15 Sample: S98T003400

Sample Name: S98T003400 DUP

Date: 01/22/1999 14:32:24

Data File : C:\DX\DATA\99012201.D16

Method : C:\DX\METHOD\KIT.MET

ACI Address: 1 System: 1 Inject#: 16

Detector: CDM-1

Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	224	1096	1	
2	1.03	fluoride	0.189	51	136	1	3.70
3	1.45	chloride	7.666	938	4507	1	1.17
4	1.71	nitrite	138.670	10410	56523	1	0.19
5	2.79	nitrate	261.468	14377	99985	1	-3.34
6	4.09	phosphate	55.618	689	8751	1	3.20
7	5.17	sulfate	129.829	4428	60424	1	3.26
8	6.67	oxalate	263.014	5108	87074	1	3.31
Totals			856.454	36224	318496		

File: 99012201.D16 Sample: S98T003400 DUP

Data Reprocessed On 01/22/1999 14:47:39

Sample Name: S98T003400 SPK Date: 01/22/1999 14:43:38
 Data File : C:\DX\DATA\99012201.D17
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 17 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	212	926	2	
2	1.01	fluoride	13.495	1565	9863	2	2.36
3	1.45	chloride	24.586	3052	14650	1	1.17
4	1.72	nitrite	269.710	20501	110336	1	0.58
5	2.50	bromide	107.937	5559	29928	1	-2.34
6	2.77	nitrate	451.431	24291	176738	1	0.12
7	4.08	phosphate	173.606	2240	27885	1	3.03
8	5.17	sulfate	143.556	4856	66919	1	3.26
9	6.72	oxalate	213.913	4533	70584	1	4.13
Totals			1398.233	66810	507829		

File: 99012201.D17 Sample: S98T003400 SPK

LABCORE Completed Worklist Report for Worklist# 27692

Analyst: adp

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0 @IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0 @IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0 @IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0 @IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0 @IC4G-QC F*4	QC	5.89e1	6.30e+01	106.961 % Recovery	
2	LCS-INST	0 @IC4G-QC ACETATE2	QC	1.24e2	1.15e+02	92.742 % Recovery	
2	LCS-INST	0 @IC4G-QC FORMATE2	QC	1.19e2	1.14e+02	95.798 % Recovery	
2	LCS-INST	0 @IC4G-QC GLYCOLTI	QC	1.07e2	1.05e+02	98.131 % Recovery	
3	CCV	0 @IC4G-QC F*4	QC	6.21e1	5.97e+01	96.135 % Recovery	
3	CCV	0 @IC4G-QC ACETATE2	QC	1.23e2	1.22e+02	99.187 % Recovery	
3	CCV	0 @IC4G-QC FORMATE2	QC	1.30e2	1.26e+02	96.923 % Recovery	
3	CCV	0 @IC4G-QC GLYCOLTI	QC	1.02e2	9.95e+01	97.549 % Recovery	
4	BLNK-PREP	0 @IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5	SAMPLE	S98T003390 0 W @IC4G-01 F*4-01	SOLID	N/A	1.49e+03	67.720 ug/g	
5	SAMPLE	S98T003390 0 W @IC4G-01 ACETATE2	SOLID	N/A	1.352e+02	43.680 ug/g	
5	SAMPLE	S98T003390 0 W @IC4G-01 FORMATE2	SOLID	N/A	< 1.354e+02	135.400 ug/g	
5	SAMPLE	S98T003390 0 W @IC4G-01 GLYCOLTI	SOLID	N/A	< 1.223e+02	122.300 ug/g	
6	DUP	S98T003390 0 W @IC4G-01 F*4-01	SOLID	1.49e+03	1.65e+03	10.191 RPD	
6	DUP	S98T003390 0 W @IC4G-01 ACETATE2	SOLID	1.35e+02	1.42e+02	5.054 RPD	
6	DUP	S98T003390 0 W @IC4G-01 FORMATE2	SOLID	<1.35e2	<1.33e2	RPD	
6	DUP	S98T003390 0 W @IC4G-01 GLYCOLTI	SOLID	<1.22e2	<1.20e2	RPD	
7	SPK	S98T003390 0 W @IC4G-01 F*4-01	SOLID	5.89e1	7.04e+01	119.525 % Recovery	
7	SPK	S98T003390 0 W @IC4G-01 ACETATE2	SOLID	1.24e2	1.13e+02	91.129 % Recovery	
7	SPK	S98T003390 0 W @IC4G-01 FORMATE2	SOLID	1.19e2	1.14e+02	95.798 % Recovery	
7	SPK	S98T003390 0 W @IC4G-01 GLYCOLTI	SOLID	1.07e2	1.02e+02	95.327 % Recovery	
8	SAMPLE	S98T003395 0 W @IC4G-01 F*4-01	SOLID	N/A	4.391e+02	69.000 ug/g	
8	SAMPLE	S98T003395 0 W @IC4G-01 ACETATE2	SOLID	N/A	6.698e+01	44.520 ug/g	
8	SAMPLE	S98T003395 0 W @IC4G-01 FORMATE2	SOLID	N/A	< 1.380e+02	138.100 ug/g	
8	SAMPLE	S98T003395 0 W @IC4G-01 GLYCOLTI	SOLID	N/A	< 1.246e+02	124.600 ug/g	
9	DUP	S98T003395 0 W @IC4G-01 F*4-01	SOLID	4.39e+02	3.92e+02	11.312 RPD	
9	DUP	S98T003395 0 W @IC4G-01 ACETATE2	SOLID	6.70e+01	5.22e+01	24.832 RPD	
9	DUP	S98T003395 0 W @IC4G-01 FORMATE2	SOLID	<1.38e2	<1.32e2	RPD	
9	DUP	S98T003395 0 W @IC4G-01 GLYCOLTI	SOLID	<1.25e2	<1.19e2	RPD	
10	SPK	S98T003395 0 W @IC4G-01 F*4-01	SOLID	5.89e1	6.40e+01	108.659 % Recovery	
10	SPK	S98T003395 0 W @IC4G-01 ACETATE2	SOLID	1.24e2	1.14e+02	91.935 % Recovery	
10	SPK	S98T003395 0 W @IC4G-01 FORMATE2	SOLID	1.19e2	1.13e+02	94.958 % Recovery	
10	SPK	S98T003395 0 W @IC4G-01 GLYCOLTI	SOLID	1.07e2	1.02e+02	95.327 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27692

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 27692

_____
Analyst Signature   Date

_____
Analyst Signature   Date

1/15/99
_____
Reviewer Signature   Date

01/11/99 11:19

Page: 1

A-0004-1

LABCORE Data Entry Template for Worklist# 27692

Analyst: ADP Instrument: IC 4552 Book# 357021CMethod: LA-533-105 Rev/Mod F-034121A

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T003390	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
6 DUP	S98T003390	0 W	@IC4G-01	SOLID		
7 SPK	S98T003390	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T003395	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
9 DUP	S98T003395	0 W	@IC4G-01	SOLID		
10 SPK	S98T003395	0 W	@IC4G-01	SOLID		

Final page for worklist # 27692

See Attached for signature
 Analyst Signature _____ Date _____

_____
 Analyst Signature _____ Date _____

Data Entry Comments:

updated 1-12-99
John Womell

validated 1/13/99 gm Lugo

27692 JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

12/28/98 14:33
A-0004-I

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 27692

Analyst: ADD Instrument: IC 4552 Book# 35N21-C LCS
34N21-A CCV
Method: LA-533-105 Rev/Mod F-0
Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T003390	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
6 DUP	S98T003390	0 W	@IC4G-01	SOLID		
7 SPK	S98T003390	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T003395	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
9 DUP	S98T003395	0 W	@IC4G-01	SOLID		
10 SPK	S98T003395	0 W	@IC4G-01	SOLID		
11 SAMPLE	S98T003400	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
12 DUP	S98T003400	0 W	@IC4G-01	SOLID		
13 SPK	S98T003400	0 W	@IC4G-01	SOLID		

Final page for worklist # 27692

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S98T003400 need to be put on another WL 27850

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 35N21-C LCS Date: 01/10/1999 02:07:00
 Data File : C:\DX\DATA\99010901.D15
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 15 Detector: CDM-2
 Analyst : *Anthony Kurits* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.90		0.000	10990	173348	1	
2	2.07		0.000	13109	247087	2	
3	2.27		0.000	3291	18578	2	
11	5.24		0.000	1857	40333	2	
13	5.83		0.000	3571	95797	2	
14	6.45		0.000	2110	26124	2	
15	7.41	fluoride	63.009 ^{107.17}	1830837	27066446	2	0.00
16	8.13	glycolate	105.122 ^{98.25}	513991	9360447	2	0.00
17	8.64	acetate	114.533 ^{92.37}	585488	11464333	2	0.00
18	9.39	formate	113.967 ^{95.07}	7006536	19137436	2	0.00
19	10.67		0.000	1357566	17703442	2	
20	11.31		0.000	6967821	209438055	2	
21	11.71		0.000	6960730	78404468	2	
22	11.97		0.000	9406670	776536284	3	
23	18.00		0.000	1380	23707	4	
Totals			396.631	28665945	1149735885		

File: 99010901.D15 Sample: 35N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1436 TO 1449.

```

=====
Sample Name: 34N21-A CCV                      Date: 01/10/1999 02:27:16
Data File  : C:\DX\DATA\99010901.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz    0.00  19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	13900	318836	1	
2	1.60		0.000	3731	25360	1	
6	2.79		0.000	3840	24980	2	
7	2.88		0.000	4780	31189	2	
8	3.03		0.000	4623	26685	2	
9	3.16		0.000	4420	24146	2	
10	3.35		0.000	3020	17921	2	
11	3.44		0.000	2600	21245	2	
12	3.60		0.000	3249	18162	2	
13	3.75		0.000	3371	28318	2	
16	4.23		0.000	2267	17676	2	
25	5.36		0.000	3664	18302	1	
26	5.68		0.000	3082	23936	2	
27	5.88		0.000	2922	43748	2	
30	6.52		0.000	2689	24650	2	
31	6.72		0.000	4368	28800	2	
32	7.39	fluoride	59.667	1722109	25589314	2	0.00
33	8.13	glycolate	99.534	489915	8860698	2	0.00
34	8.64	acetate	122.341	621694	12227512	2	0.00
35	9.36	formate	126.220	1100576	21234841	2	0.00
36	10.67		0.000	100426898	30069937	2	
37	11.33		0.000	6785315	210845054	2	
38	11.71		0.000	6447789	77214861	2	
39	12.03		0.000	8943910	719961812	3	
41	18.51		0.000	2346	26389	4	
Totals			407.762	126607077	1106724373		

```

=====
Sample Name: BLANK                      Date: 01/10/1999 01:06:07
Data File  : C:\DX\DATA\99010901.D14
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14      Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1    5700  5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	7539	129990	2	
2	0.67		0.000	4009	61603	2	
3	1.07		0.000	2059	17583	3	
7	2.73		0.000	4621	31462	2	
8	2.80		0.000	4674	45447	2	
9	2.99		0.000	4470	46114	2	
10	3.21		0.000	2873	37218	2	
22	5.11		0.000	1966	25222	2	
23	5.48		0.000	1931	29833	2	
26	6.15		0.000	2067	18892	2	
31	7.23		0.000	2248	45208	2	
34	8.43		0.000	1820	36132	1	
35	9.20		0.000	2919	23528	2	
36	9.44	formate	0.006	3116	55385	2	0.00
37	11.33		0.000	7136720	244915714	2	
38	11.71		0.000	6718509	90973263	2	
39	12.03		0.000	9240042	711487183	3	
40	18.16		0.000	2258	20825	4	
41	18.37		0.000	2286	38136	4	
42	18.67		0.000	2514	19394	4	
Totals			0.006	23148642	1048058132		

File: 99010901.D14 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 01/10/1999 02:47:26
Data File  : C:\DX\DATA\99010901.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17         Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700  5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	10790	162536	2	
2	1.00		0.000	1428	20215	2	
3	1.53		0.000	1720	16464	1	
15	3.57		0.000	2412	18984	2	
37	6.77		0.000	1830	24896	2	
38	7.15		0.000	2880	21740	2	
39	7.36	fluoride	0.009	2700	37333	2	0.00
40	7.63		0.000	2170	32774	2	
41	8.05	glycolate	-0.016	1550	29662	2	0.00
45	9.39	formate	0.004	1459	20181	2	0.00
46	9.63		0.000	1305	29086	2	
47	11.31		0.000	6881403	225235631	2	
48	11.73		0.000	6389678	86030042	2	
49	12.03		0.000	8614242	700974056	3	
Totals			-0.003	21915566	1012653599		

File: 99010901.D17 Sample: PREP BLANK

```

=====
Sample Name: S98T003390 SAM                      Date: 01/10/1999 03:07:52
Data File  : C:\DX\DATA\99010901.D18
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 18              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 11 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	9283	143442	1	
3	1.70		0.000	54740	703856	1	
4	2.95		0.000	12945	303112	2	
5	3.17		0.000	11109	105358	2	
6	3.33		0.000	7998	31790	2	
7	3.41		0.000	7652	47960	2	
8	3.53		0.000	8783	85290	2	
9	3.77		0.000	7546	68748	2	
10	3.93		0.000	5375	20458	2	
11	4.03		0.000	5608	24995	2	
12	4.11		0.000	5542	38265	2	
13	4.29		0.000	4609	40181	2	
19	4.97		0.000	4840	48362	2	
24	5.95		0.000	3126	42548	2	
27	7.33	fluoride	7.517	1962444	29730326	2	0.00
28	8.56	acetate	0.681	39525	839344	2	0.00
29	9.28	formate	0.582	41667	838605	2	0.00
30	10.93		0.000	2780708	42594678	2	
31	11.20		0.000	12311547	218230794	2	
32	11.71		0.000	330308042	5531232931	3	
Totals			8.780	347593087	5825171043		

File: 99010901.D18 Sample: S98T003390 SAM

HNF-1663 REV. 0

```

=====
Sample Name: S98T003390 DUP                      Date: 01/10/1999 03:29:52
Data File  : C:\DX\DATA\99010901.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          11    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	8701	156106	2	
3	1.70		0.000	57515	754161	2	
4	3.04		0.000	5225	110819	2	
5	3.12		0.000	4696	24485	3	
23	5.91		0.000	1795	32272	2	
25	7.33	fluoride	8.481	2214634	33667893	3	0.00
26	8.03	glycolate	-0.107	7090	84926	4	0.00
27	8.56	acetate	0.728	40834	881872	2	0.00
28	9.28	formate	0.538	38418	772538	2	0.00
29	10.93		0.000	2795080	38546542	2	
30	11.20		0.000	12273914	230958597	2	
31	11.71		0.000	341110397	5727748753	2	
Totals			9.640	358558299	6033738965		

File: 99010901.D19 Sample: S98T003390 DUP

```

=====
Sample Name: S98T003390 SPK                      Date: 01/10/1999 03:56:02
Data File  : C:\DX\DATA\99010901.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          11 5700 5Hz 0.00 19.00 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
5	1.70		0.000	68820	804616	1	
6	2.91		0.000	8419	274821	2	
7	3.40		0.000	7401	73945	2	
8	3.52		0.000	6737	50828	2	
9	3.69		0.000	5383	26048	2	
10	3.80		0.000	5602	48318	2	
11	3.93		0.000	5596	66291	2	
12	4.19		0.000	3726	33411	2	
13	4.39		0.000	3146	27707	2	
14	4.57		0.000	3030	71325	2	
15	5.03		0.000	2540	27946	2	
18	5.93		0.000	2011	30623	2	
19	7.31	fluoride	14.556	3847841	58948560	2	0.00
20	8.03	glycolate	10.206	446639	8341446	2	0.00
21	8.53	acetate	11.945	564633	10989739	2	0.00
22	9.25	formate	11.396	919330	17544970	2	0.00
23	10.05		0.000	2555	24192	1	
24	11.20		0.000	12481879	296920315	2	
25	11.71		0.000	337717971	5655310317	3	
26	18.24		0.000	2260	40683	4	
27	18.67		0.000	2140	35097	4	
Totals			48.104	356107659	6049691197		

```

=====
Sample Name: S98T003395 SAM                      Date: 01/10/1999 04:15:58
Data File  : C:\DX\DATA\99010901.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 21              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      11      5700 5Hz      0.00 19.00      16000
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	9563	169184	1	
3	1.73		0.000	60984	746600	1	
8	2.96		0.000	3941	45173	2	
10	3.36		0.000	3313	16284	2	
24	5.36		0.000	3050	26964	2	
25	5.57		0.000	3410	19887	2	
26	5.75		0.000	3119	31514	2	
27	5.93		0.000	2312	36511	2	
30	7.31	fluoride	2.170	506417	8223161	3	0.00
31	7.97		0.000	3779	76677	4	
32	8.56	acetate	0.331	25260	522756	2	0.00
33	9.28	formate	0.468	31363	665996	2	0.00
35	11.23		0.000	10018477	243090230	2	
36	11.71		0.000	344917566	5640414759	2	
Totals			2.970	355592555	5894085697		

File: 99010901.D21 Sample: S98T003395 SAM

Data Reprocessed On 01/11/1999 14:55:20

```

=====
Sample Name: S98T003395 DUP                      Date: 01/10/1999 04:35:46
Data File  : E:\DATA\99010901.D22
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 22              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          11    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	2380	30180	1	
5	1.70		0.000	65500	762912	1	
6	2.75		0.000	6889	75150	2	
7	2.83		0.000	6221	50774	2	
8	3.00		0.000	5343	39451	2	
16	4.44		0.000	2484	20408	2	
17	4.64		0.000	1971	21341	2	
19	5.15		0.000	2502	18222	2	
22	5.60		0.000	2790	21219	2	
26	6.72		0.000	2777	20318	2	
27	7.28	fluoride	2.022	473815	7635974	2	0.00
28	8.51	acetate	0.269	22422	466249	2	0.00
29	9.25	formate	0.423	30999	596555	2	0.00
31	11.23		0.000	9897612	240856525	2	
32	11.71		0.000	357116755	5949774164	2	
Totals			2.714	367640460	6200389441		

File: 99010901.D22 Sample: S98T003395 DUP

```

=====
Sample Name: S98T003395 SPK                      Date: 01/10/1999 04:56:03
Data File  : C:\DX\DATA\99010901.D23
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 23              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          11    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	6017	159042	1	
3	1.70		0.000	57600	719436	1	
4	2.84		0.000	1646	29760	3	
11	3.87		0.000	3020	23412	2	
27	6.12		0.000	2006	17406	2	
30	7.31	fluoride	8.574	2173694	34050418	2	0.00
31	8.05	glycolate	10.167	443855	8309037	2	0.00
32	8.56	acetate	11.686	544060	10757112	2	0.00
33	9.28	formate	11.257	892625	17327579	2	0.00
35	11.23		0.000	10093650	244539256	2	
36	11.71		0.000	345312634	5709082976	2	
Totals			41.684	359530807	6025015433		

File: 99010901.D23 Sample: S98T003395 SPK

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 5 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 1450 THRU 1838)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	ION CHROMATOGRAPHIC ANALYSIS, IC WORKLIST 27693 THRU 28718		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

HNF-1663 REV. 0

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635
Specific Gravity Analysis (SpG)	
SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684
Hydroxide Analysis (OH)	
OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775
Ammonia Analysis (NH3)	
NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992
Chromium VI Analysis (CRVI)	
CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)	
GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)	
Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)	
Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 27693

Analyst: adp

Instrument: IC45S2

Book#: 35N21A

Method: LA-533-105

Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0	@IC4G-QC F*4	QC	6.05e1	6.12e+01	101.157 %	Recovery
2	LCS-INST	0	@IC4G-QC ACETATE2	QC	1.19e2	9.00e+01	75.630 %	Recovery
2	LCS-INST	0	@IC4G-QC FORMATE2	QC	1.28e2	1.17e+02	91.406 %	Recovery
2	LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.16e2	1.08e+02	93.103 %	Recovery
3	CCV	0	@IC4G-QC F*4	QC	6.21e1	5.81e+01	93.559 %	Recovery
3	CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.15e+02	93.496 %	Recovery
3	CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.21e+02	93.077 %	Recovery
3	CCV	0	@IC4G-QC GLYCOLT1	QC	1.02e2	9.49e+01	93.039 %	Recovery
4	BLNK-PREP	0	@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4	BLNK-PREP	0	@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5	SAMPLE	S98T002714	0 W	@IC4G-01 ACETATE2	SOLID	N/A	8.025e+02	82.360 ug/g
5	SAMPLE	S98T002714	0 W	@IC4G-01 FORMATE2	SOLID	N/A	2.783e+02	255.400 ug/g
5	SAMPLE	S98T002714	0 W	@IC4G-01 GLYCOLT1	SOLID	N/A	< 2.306e+02	230.600 ug/g
6	DUP	S98T002714	0 W	@IC4G-01 F*4-01	SOLID	1.35e+04	8.68e+03	43.463 RPD
6	DUP	S98T002714	0 W	@IC4G-01 ACETATE2	SOLID	8.02e+02	6.88e+02	15.302 RPD
6	DUP	S98T002714	0 W	@IC4G-01 FORMATE2	SOLID	2.78e+02	2.53e+02	9.416 RPD
6	DUP	S98T002714	0 W	@IC4G-01 GLYCOLT1	SOLID	<2.31e2	<2.28e2	RPD
7	SAMPLE	S98T002741	0 W	@IC4G-01 F*4-01	SOLID	N/A	3.042e+03	130.400 ug/g
7	SAMPLE	S98T002741	0 W	@IC4G-01 FORMATE2	SOLID	N/A	< 2.608e+02	260.800 ug/g
7	SAMPLE	S98T002741	0 W	@IC4G-01 GLYCOLT1	SOLID	N/A	< 2.355e+02	235.600 ug/g
8	DUP	S98T002741	0 W	@IC4G-01 F*4-01	SOLID	3.04e+03	2.79e+03	8.576 RPD
8	DUP	S98T002741	0 W	@IC4G-01 ACETATE2	SOLID	2.26e+02	1.70e+02	28.283 RPD
8	DUP	S98T002741	0 W	@IC4G-01 FORMATE2	SOLID	<2.61e2	<2.63e2	RPD
8	DUP	S98T002741	0 W	@IC4G-01 GLYCOLT1	SOLID	<2.36e2	<2.38e2	RPD
9	SPK	S98T002741	0 W	@IC4G-01 F*4-01	SOLID	6.05e1	7.31e+01	120.826 % Recovery
9	SPK	S98T002741	0 W	@IC4G-01 ACETATE2	SOLID	1.19e2	9.75e+01	81.933 % Recovery
9	SPK	S98T002741	0 W	@IC4G-01 FORMATE2	SOLID	1.28e2	1.25e+02	97.656 % Recovery
9	SPK	S98T002741	0 W	@IC4G-01 GLYCOLT1	SOLID	1.16e2	1.12e+02	96.552 % Recovery
10	SAMPLE	S98T002848	0 W	@IC4G-01 FORMATE2	SOLID	N/A	< 2.549e+02	254.900 ug/g
11	DUP	S98T002848	0 W	@IC4G-01 F*4-01	SOLID	8.52e+02	1.60e+03	61.011 RPD
11	DUP	S98T002848	0 W	@IC4G-01 ACETATE2	SOLID	2.96e+02	3.52e+02	17.284 RPD
11	DUP	S98T002848	0 W	@IC4G-01 FORMATE2	SOLID	<2.55e2	<2.55e2	RPD
11	DUP	S98T002848	0 W	@IC4G-01 GLYCOLT1	SOLID	<2.30e2	<2.30e2	RPD
12	SPK	S98T002848	0 W	@IC4G-01 F*4-01	SOLID	6.05e1	6.10e+01	100.826 % Recovery
12	SPK	S98T002848	0 W	@IC4G-01 ACETATE2	SOLID	1.19e2	6.39e+01	53.697 % Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27693

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
12	SPK	S98T002848	0	W	@IC40-01 FORMATE2	SOLID	1.28e2	1.07e+02	83.594 %	Recovery
12	SPK	S98T002848	0	W	@IC40-01 GLYCOLT1	SOLID	1.16e2	8.30e+01	71.552 %	Recovery

Final page for worklist# 27693

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 1/25/99

12/28/98 14:34
A-0004-1

LABCORE Data Entry Template for Worklist# 27693

Analyst: ADD Instrument: IC 4582 Book# 35N21-A LCS
 Method: LA-533-105 Rev/Mod F-0 34N21-A ccv

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	CCB		@IC4G-QC	QC		
2	LCS-INST		@IC4G-QC	QC		
3	CCV		@IC4G-QC	QC		
4	BLNK-PREP		@IC4G-01	SOLID		
5	SAMPLE	S98T002714 0 W	@IC4G-01 SOLID		98000436 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
6	DUP	S98T002714 0 W	@IC4G-01 SOLID			
7	SAMPLE	S98T002741 0 W	@IC4G-01 SOLID		98000490 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
8	DUP	S98T002741 0 W	@IC4G-01 SOLID			
9	SPK	S98T002741 0 W	@IC4G-01 SOLID			
10	SAMPLE	S98T002848 0 W	@IC4G-01 SOLID		98000437 BY-105 (2)	
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
11	DUP	S98T002848 0 W	@IC4G-01 SOLID			
12	SPK	S98T002848 0 W	@IC4G-01 SOLID			

Final page for worklist # 27693

Anthony Puro ¹⁴⁴³ 01-9-21-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 1-21-99John Howell27693JANCSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

598T002741/4 2714 reanalyzed

on WL# 28347.

598T002848 reanalyzed on

WL# 28406. for 7/24/99Resun 598T002741 acetate,598T002714 fluoride598T002848 F, acetate, glycolatefor HighRPDx poor spk.Rest validated. Janv Juy1/21/99

Sample Name: 35N21-A LCS Date: 01/20/1999 21:49:21
 Data File : C:\DX\DATA\99012001.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
 Analyst: *Anthony P. Smith* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	23110	715924	1	
2	2.10		0.000	37203	573202	2	
3	2.27		0.000	38647	200196	2	
4	2.37		0.000	37455	214298	2	
5	2.51		0.000	35814	202491	2	
6	2.60		0.000	36676	231076	2	
7	2.73		0.000	34735	482185	2	
8	3.04		0.000	31051	421901	2	
9	3.35		0.000	30247	399293	2	
10	3.48		0.000	26627	151698	2	
11	3.63		0.000	25852	294704	2	
12	3.81		0.000	27215	108319	2	
13	3.89		0.000	28070	242069	2	
14	4.04		0.000	25496	125722	2	
15	4.13		0.000	25997	267852	2	
16	4.36		0.000	21118	372246	2	
17	4.69		0.000	16546	249557	2	
18	5.00		0.000	9782	53448	2	
20	5.41		0.000	11591	112155	2	
21	5.57		0.000	9932	59425	2	
22	5.69		0.000	8107	101836	2	
26	7.20	fluoride	61.205	1768339	26268807	2	0.00
27	7.89	glycolate	108.100	525441	9627296	2	0.00
28	8.37	acetate	90.050	466073	9068480	2	0.00
29	9.09	formate	117.147	1000821	19681227	2	0.00
Totals			376.502	4301944	70225408		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1453 TO 1465.

Sample Name: 34N21-A CCV Date: 01/20/1999 22:10:09
 Data File : C:\DX\DATA\99012001.D04
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	25625	757388	2	
9	3.04		0.000	4907	76859	2	
17	4.60		0.000	5347	126978	2	
19	5.51		0.000	7490	77991	2	
21	5.80		0.000	6993	71000	2	
22	6.00		0.000	3979	60986	2	
24	7.20	fluoride	58.118	1669262	24905191	2	0.00
25	7.89	glycolate	94.865	467622	8444126	2	0.00
26	8.37	acetate	114.558	582387	11466777	2	0.00
27	9.09	formate	120.842	1040334	20313421	2	0.00
Totals			388.383	3813945	66300717		

File: 99012001.D04 Sample: 34N21-A CCV

Sample Name: BLANK

Date: 01/20/1999 21:28:59

Data File : C:\DX\DATA\99012001.D02

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 2

Detector: CDM-2

Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	2.21		0.000	110238	7426900	1	
2	5.57		0.000	118800	3086196	2	
3	5.68		0.000	122270	14869270	2	
4	9.63		0.000	1533	492501	2	
Totals			0.000	352842	25874867		

File: 99012001.D02 Sample: BLANK

=====

Sample Name: PREP BLANK	Date: 01/20/1999 22:30:26
Data File : C:\DX\DATA\99012001.D05	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 5	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	11607	208432	1	
3	1.27		0.000	3385	64816	1	
9	3.07		0.000	4308	67348	2	
10	3.39		0.000	5715	143739	2	
14	4.61		0.000	4726	77597	2	
27	6.75		0.000	5709	125099	2	
29	7.68		0.000	4256	106768	2	
32	9.12	formate	0.006	3934	51865	2	0.00
Totals			0.006	43639	845663		

File: 99012001.D05 Sample: PREP BLANK

```

=====
Sample Name: S98T002714 SAM                      Date: 01/20/1999 23:14:20
Data File  : C:\DX\DATA\99012001.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 5700 5Hz 0.00 19.00 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
6	2.19		0.000	3880	68426	1	
21	4.64		0.000	5596	56987	2	
22	4.97		0.000	8552	92421	2	
23	5.11		0.000	8258	148236	2	
25	5.55		0.000	10198	166223	2	
26	5.87		0.000	8073	101860	2	
27	7.17	fluoride	68.641	10089623	155398853	2	0.00
28	8.32	acetate	4.092	97828	2160765	2	0.00
29	9.04	formate	1.419	55259	1085011	2	0.00
30	9.84		0.000	15167	270982	1	
Totals			74.152	10302435	159549764		

File: 99012001.D07 Sample: S98T002714 SAM

Sample Name: S98T002714 DUP Date: 01/20/1999 23:34:53
 Data File : C:\DX\DATA\99012001.D08
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 8 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	2481	67792	1	
3	1.70		0.000	13214	114516	1	
4	2.13		0.000	5262	84640	1	
7	2.81		0.000	5437	77235	2	
17	4.48		0.000	5546	78801	2	
23	5.53		0.000	3201	96856	2	
24	7.15	fluoride	44.869	6404565	97743799	2	0.00
25	8.29	acetate	3.557	87985	1907633	2	0.00
26	9.04	formate	1.307	50871	995484	2	0.00
27	9.81		0.000	13778	282923	2	

Totals 49.732 6592340 101449680

File: 99012001.D08 Sample: S98T002714 DUP

```

=====
Sample Name: S98T002741 SAM                      Date: 01/20/1999 23:55:34
Data File  : C:\DX\DATA\99012001.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz    0.00 19.00    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	8239	494480	2	
3	2.21		0.000	1103	87918	2	
4	2.69		0.000	533	57276	2	
24	5.84		0.000	6294	66739	2	
27	7.17	fluoride	15.187	2124943	31517205	2	0.00
28	8.32	acetate	1.131	32870	758660	2	0.00
29	9.07	formate	1.018	37903	764684	2	0.00
31	9.79		0.000	4713	73019	2	
Totals			17.335	2216597	33819981		

File: 99012001.D09 Sample: S98T002741 SAM

Sample Name: S98T002741 DUP Date: 01/21/1999 00:17:11
 Data File : C:\DX\DATA\99012001.D10
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 10 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	10560	285604	1	
2	1.17		0.000	3540	50544	1	
30	7.17	fluoride	13.830	1944784	28619726	2	0.00
31	8.32	acetate	0.844	30445	623016	2	0.00
32	9.07	formate	0.974	36404	729511	2	0.00
33	9.87		0.000	5900	123802	1	
Totals			15.648	2031633	30432203		

File: 99012001.D10 Sample: S98T002741 DUP

```

=====
Sample Name: S98T002741 SPK                      Date: 01/21/1999 00:38:48
Data File  : C:\DX\DATA\99012001.D11
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 11              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz   0.00 19.00   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	10124	391128	1	
2	1.70		0.000	10113	105490	1	
5	2.85		0.000	9513	159980	2	
6	3.09		0.000	7696	75438	2	
26	7.17	fluoride	29.800	4288628	63416239	2	0.00
27	7.84	glycolate	22.386	522069	9588606	2	0.00
28	8.32	acetate	20.624	509186	9963463	2	0.00
29	9.04	formate	24.996	1052862	20206736	3	0.00
Totals			97.807	6410192	103907079		

File: 99012001.D11 Sample: S98T002741 SPK

```

=====
Sample Name: S98T002748 SAM                      Date: 01/21/1999 00:58:42
Data File  : C:\DX\DATA\99012001.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz   0.00 19.00    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	10067	98380	1	
10	3.04		0.000	5363	90298	2	
13	4.08		0.000	4839	64477	2	
19	5.05		0.000	6033	127385	2	
20	5.47		0.000	5431	62257	2	
22	5.79		0.000	4800	68766	2	
26	7.15	fluoride	4.354	566653	8662856	2	0.00
28	8.32	acetate	1.510	49710	938127	2	0.00
29	9.04	formate	1.003	42670	753031	2	0.00
31	9.81		0.000	8046	130271	2	
Totals			6.867	703611	10995847		

File: 99012001.D12 Sample: S98T002748 SAM

```

=====
Sample Name: S98T002748  DUP                      Date: 01/21/1999 01:19:37
Data File  : C:\DX\DATA\99012001.D13
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 13                Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz   0.00 19.00   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	6669	296224	1	
7	2.99		0.000	6168	53875	2	
9	3.27		0.000	8103	106553	2	
15	4.37		0.000	4327	62005	2	
21	5.36		0.000	6805	61709	2	
22	5.51		0.000	6860	64613	2	
24	5.80		0.000	6231	60870	2	
28	7.15	fluoride	8.194	1115406	16690221	3	0.00
29	7.87	glycolate	-0.227	6098	75764	4	0.00
30	8.32	acetate	1.797	53645	1074357	2	0.00
31	9.07	formate	1.273	48072	968685	2	0.00
32	9.89		0.000	10703	223450	1	
Totals			11.036	1279086	19738326		

File: 99012001.D13 Sample: S98T002748 DUP

```

=====
Sample Name: S98T002748 SPK                      Date: 01/21/1999 01:50:01
Data File  : C:\DX\DATA\99012001.D14
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          21    5700 5Hz    0.00 19.00    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	4986	85337	2	
2	1.73		0.000	12434	102850	2	
3	2.64		0.000	7099	59908	2	
5	2.92		0.000	8091	101719	2	
6	3.16		0.000	6330	58784	2	
11	3.95		0.000	4462	60636	2	
26	7.12	fluoride	16.545	2370200	34429024	1	0.00
27	7.81	glycolate	16.605	444772	7111547	2	0.00
28	8.29	acetate	14.291	425272	6979483	2	0.00
29	8.99	formate	21.431	966892	17278180	1	0.00
Totals			68.872	4250539	66267467		

File: 99012001.D14 Sample: S98T002748 SPK

LABCORE Completed Worklist Report for Worklist# 27732

Analyst: adp Instrument: IC40S1 Book#: 51N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1	CCB	0	@IC-QC	F	QC	1	<1.20e-2	ug/mL	
1	CCB	0	@IC-QC	CL	QC	1	<1.70e-2	ug/mL	
1	CCB	0	@IC-QC	NO2	QC	1	<1.08e-1	ug/mL	
1	CCB	0	@IC-QC	BR	QC	1	<1.25e-1	ug/mL	
1	CCB	0	@IC-QC	NO3	QC	1	<1.39e-1	ug/mL	
1	CCB	0	@IC-QC	P04	QC	1	<1.20e-1	ug/mL	
1	CCB	0	@IC-QC	SO4	QC	1	<1.38e-1	ug/mL	
1	CCB	0	@IC-QC	OxALATE2	QC	1	<1.05e-1	ug/mL	
2	LCS-INST	0	@IC-QC	F	QC	5.92e1	6.43e+01	108.615 % Recovery	
2	LCS-INST	0	@IC-QC	CL	QC	7.96e1	8.32e+01	104.523 % Recovery	
2	LCS-INST	0	@IC-QC	NO2	QC	5.15e2	5.21e+02	101.165 % Recovery	
2	LCS-INST	0	@IC-QC	BR	QC	5.86e2	5.85e+02	99.829 % Recovery	
2	LCS-INST	0	@IC-QC	NO3	QC	5.92e2	5.88e+02	99.324 % Recovery	
2	LCS-INST	0	@IC-QC	P04	QC	5.48e2	5.70e+02	104.015 % Recovery	
2	LCS-INST	0	@IC-QC	SO4	QC	6.38e2	6.48e+02	101.567 % Recovery	
2	LCS-INST	0	@IC-QC	OxALATE2	QC	5.22e2	5.71e+02	109.387 % Recovery	
3	CCV	0	@IC-QC	F	QC	6.40e1	6.59e+01	102.969 % Recovery	
3	CCV	0	@IC-QC	CL	QC	8.95e1	9.26e+01	103.464 % Recovery	
3	CCV	0	@IC-QC	NO2	QC	5.45e2	5.56e+02	102.018 % Recovery	
3	CCV	0	@IC-QC	BR	QC	6.30e2	6.43e+02	102.063 % Recovery	
3	CCV	0	@IC-QC	NO3	QC	7.00e2	7.19e+02	102.714 % Recovery	
3	CCV	0	@IC-QC	P04	QC	6.35e2	6.67e+02	105.039 % Recovery	
3	CCV	0	@IC-QC	SO4	QC	7.00e2	7.04e+02	100.571 % Recovery	
3	CCV	0	@IC-QC	OxALATE2	QC	5.27e2	5.69e+02	107.970 % Recovery	
4	BLNK-PREP	0	@IC-01	F-02	SOLID	1	<1.20e-2	ug/g	
4	BLNK-PREP	0	@IC-01	CL-02	SOLID	1	<1.70e-2	ug/g	
4	BLNK-PREP	0	@IC-01	NO2-02	SOLID	1	5.04e-01	0.504 ug/g	
4	BLNK-PREP	0	@IC-01	BR-02	SOLID	1	<1.25e-1	ug/g	
4	BLNK-PREP	0	@IC-01	NO3-02	SOLID	1	1.50e-01	0.150 ug/g	
4	BLNK-PREP	0	@IC-01	P04-02	SOLID	1	<1.20e-1	ug/g	
4	BLNK-PREP	0	@IC-01	SO4-02	SOLID	1	<1.38e-1	ug/g	
4	BLNK-PREP	0	@IC-01	OxALATE2	SOLID	1	<1.05e-1	ug/g	
5	SAMPLE	S98T002612	0 W	@IC-01	NO2-02	SOLID	N/A	4.891e+03	945.000 ug/g
5	SAMPLE	S98T002612	0 W	@IC-01	BR-02	SOLID	N/A	9.779e+02	977.900 ug/g
5	SAMPLE	S98T002612	0 W	@IC-01	NO3-02	SOLID	N/A	2.995e+05	1087.000 ug/g
5	SAMPLE	S98T002612	0 W	@IC-01	SO4-02	SOLID	N/A	1.170e+04	1080.000 ug/g
6	DUP	S98T002612	0 W	@IC-01	F-02	SOLID	?	8.98e+03	RPD
6	DUP	S98T002612	0 W	@IC-01	CL-02	SOLID	?	4.99e+02	RPD
6	DUP	S98T002612	0 W	@IC-01	NO2-02	SOLID	4.89e+03	7.11e+03	37.000 RPD
6	DUP	S98T002612	0 W	@IC-01	BR-02	SOLID	<9.78e2	<9.80e2	RPD
6	DUP	S98T002612	0 W	@IC-01	NO3-02	SOLID	3.00e+05	1.88e+05	45.902 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27732

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002612	0 W	@IC-01	P04-02	SOLID	?	1.64e+04	RFD
6 DUP	S98T002612	0 W	@IC-01	S04-02	SOLID	1.17e+04	1.54e+04	27.306 RFD
6 DUP	S98T002612	0 W	@IC-01	OXALATE2	SOLID	?	1.44e+04	RFD
7 SAMPLE	S98T002614	0 W	@IC-01	F-02	SOLID	N/A	9.602e+01	96.020 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	CL-02	SOLID	N/A	2.091e+03	136.000 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	N02-02	SOLID	N/A	3.478e+04	864.300 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	BR-02	SOLID	N/A	1.000e+03	1000.000 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	N03-02	SOLID	N/A	7.553e+04	1112.000 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	P04-02	SOLID	N/A	5.595e+03	960.200 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	S04-02	SOLID	N/A	3.774e+03	1104.000 ug/g
7 SAMPLE	S98T002614	0 W	@IC-01	OXALATE2	SOLID	N/A	8.709e+03	840.100 ug/g
8 DUP	S98T002614	0 W	@IC-01	F-02	SOLID	<9.60e1	1.37e+02	RFD
8 DUP	S98T002614	0 W	@IC-01	CL-02	SOLID	2.09e+03	2.07e+03	0.962 RFD
8 DUP	S98T002614	0 W	@IC-01	N02-02	SOLID	3.48e+04	3.61e+04	3.667 RFD
8 DUP	S98T002614	0 W	@IC-01	BR-02	SOLID	<1.00e3	<1.03e3	RFD
8 DUP	S98T002614	0 W	@IC-01	N03-02	SOLID	7.55e+04	7.24e+04	4.192 RFD
8 DUP	S98T002614	0 W	@IC-01	P04-02	SOLID	5.60e+03	7.05e+03	22.925 RFD
8 DUP	S98T002614	0 W	@IC-01	S04-02	SOLID	3.77e+03	3.48e+03	8.000 RFD
8 DUP	S98T002614	0 W	@IC-01	OXALATE2	SOLID	8.71e+03	7.77e+03	11.408 RFD
9 SPK	S98T002614	0 W	@IC-01	F-02	SOLID	5.92e1	7.05e+01	119.088 % Recovery
9 SPK	S98T002614	0 W	@IC-01	CL-02	SOLID	7.96e1	7.88e+01	98.995 % Recovery
9 SPK	S98T002614	0 W	@IC-01	N02-02	SOLID	5.15e2	5.50e+02	106.796 % Recovery
9 SPK	S98T002614	0 W	@IC-01	BR-02	SOLID	5.86e2	5.74e+02	97.952 % Recovery
9 SPK	S98T002614	0 W	@IC-01	N03-02	SOLID	5.92e2	6.04e+02	102.027 % Recovery
9 SPK	S98T002614	0 W	@IC-01	P04-02	SOLID	5.49e2	5.66e+02	103.285 % Recovery
9 SPK	S98T002614	0 W	@IC-01	S04-02	SOLID	6.38e2	6.36e+02	99.687 % Recovery
9 SPK	S98T002614	0 W	@IC-01	OXALATE2	SOLID	5.22e2	5.59e+02	107.088 % Recovery

Final page for worklist# 27732

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jam M. Luge
Reviewer Signature _____ Date 1/24/99

12/30/98 14:51
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27732

Analyst: ADP Instrument: IC Book# 51N21-B LCS
 Method: LA-533-105 Rev/Mod F-0 50N21-B CCV
 Worklist Comment: BY-105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T002612 0 W			@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested:	BR-02			, NO2-02		, NO3-02 , SO4-02
6	DUP	S98T002612 0 W			@IC-01	SOLID		
7	SAMPLE	S98T002614 0 W			@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested:	BR-02			, CL-02 , F-02		, NO2-02 , NO3-02 ,
			OXALATE2,			PO4-02		, SO4-02
8	DUP	S98T002614 0 W			@IC-01	SOLID		
9	SPK	S98T002614 0 W			@IC-01	SOLID		

Final page for worklist # 27732

Anthony Perreault
 Analyst Signature Date 1-22-99

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 1-22-99
John Howell
27732 JAN. CSV

validated RPD's still high
on 2/6/2 JM Faye 1/25/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 51N21-B LCS Date: 01/21/1999 23:07:05
 Data File : C:\DX\DATA\99012101.D12
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: A System: 1 Inject#: 12 Detector: CDM-1
 Analyst : *Anthony Perreault* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.86		0.000	81	325	2	
2	1.01	fluoride	64.348	1893	9777	2	1.68
3	1.46	chloride	83.174	2308	10245	1	2.10
4	1.73	nitrite	521.126	8754	44129	1	0.97
5	2.53	bromide	584.736	6077	33772	1	-1.30
6	2.85	nitrate	587.910	6978	45991	1	-1.50
7	4.13	phosphate	570.061	1506	18853	1	4.21
8	5.28	sulfate	647.557	5420	62698	1	5.39
9	6.83	oxalate	570.509	2419	38892	1	5.79
Totals			3629.421	35437	264682		

File: 99012101.D12 Sample: 51N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1468 TO 1477.

```

=====
Sample Name: 50N21-B   CCV                      Date: 01/21/1999 23:30:50
Data File  : C:\DX\DATA\99012101.D14
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 14             Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	85	330	2	
2	1.02	fluoride	65.908	1900	10018	2	3.03
3	1.48	chloride	92.566	2600	11419	1	3.50
4	1.75	nitrite	556.183	9302	47107	1	2.14
5	2.55	bromide	643.422	6506	37223	1	-0.52
6	2.87	nitrate	718.969	8245	56419	1	-0.81
7	4.16	phosphate	667.391	1758	22150	1	5.05
8	5.28	sulfate	703.762	4802	68233	1	5.39
9	6.88	oxalate	568.991	2482	38788	1	6.61
Totals			4017.192	37681	291688		

File: 99012101.D14 Sample: 50N21-B CCV

```
=====
Sample Name: BLANK                      Date: 01/21/1999 23:18:59
Data File  : C:\DX\DATA\99012101.D13
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 13    Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====
```

Calibration Volume Dilution Points Rate Start Stop Area Reject

-----
External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta

-----
Totals 0.000 0 0

File: 99012101.D13 Sample: BLANK

```

=====
Sample Name: PREP BLANK      Date: 01/21/1999 23:42:31
Data File  : C:\DX\DATA\99012101.D15
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 15      Detector: CDM-1
Analyst    :                  Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate Start Stop Area Reject
-----
External           1           1 3000 5Hz 0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.31		0.000	54	520	1	
2	1.75	nitrite	0.504	779	4299	1	2.53
3	2.92	nitrate	0.150	192	1170	1	1.04
Totals			0.654	1025	5988		

File: 99012101.D15 Sample: PREP BLANK

```

=====
Sample Name: S98T002612 SAM                      Date: 01/21/1999 23:54:28
Data File  : C:\DX\DATA\99012101.D16
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 16              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	149	588	2	
2	1.02	fluoride	39.771	2722	15013	2	3.03
3	1.47	chloride	1.632	112	489	1	3.03
4	1.74	nitrite	25.635	1039	5334	1	1.75
5	2.79	nitrate	1569.648	41047	330531	1	-3.34
6	4.16	phosphate	85.990	550	6916	1	5.05
7	5.33	sulfate	61.301	1226	14450	1	6.45
8	6.88	oxalate	63.891	645	10669	1	6.61
Totals			1847.868	47491	383989		

File: 99012101.D16 Sample: S98T002612 SAM

```

=====
Sample Name: S98T002612  DUP                      Date: 01/22/1999 00:06:19
Data File  : C:\DX\DATA\99012101.D17
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 17                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	172	705	2	
2	1.02	fluoride	46.940	3191	17799	2	3.03
3	1.47	chloride	2.609	178	782	1	3.03
4	1.75	nitrite	37.168	1480	7734	1	2.14
5	2.56	bromide	0.633	19	89	1	0.00
6	2.83	nitrate	985.357	26177	198963	1	-2.19
7	4.17	phosphate	85.736	532	6896	1	5.22
8	5.33	sulfate	80.265	1598	18941	1	6.45
9	6.88	oxalate	75.367	719	12590	1	6.61
Totals			1314.075	34065	264499		

File: 99012101.D17 Sample: S98T002612 DUP

```

=====
Sample Name: S98T002614 SAM                      Date: 01/22/1999 00:17:15
Data File  : C:\DX\DATA\99012101.D18
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 18              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000  5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	83	392	1	
2	1.04	fluoride	0.478	49	176	1	5.05
3	1.47	chloride	10.716	708	3221	1	3.03
4	1.75	nitrite	178.215	7151	37159	1	2.14
5	2.87	nitrate	387.011	10982	75237	1	-0.81
6	4.18	phosphate	28.670	191	2295	1	5.56
7	5.33	sulfate	19.337	364	4547	1	6.45
8	6.93	oxalate	44.627	482	7448	1	7.44
Totals			669.054	20011	130474		

File: 99012101.D18 Sample: S98T002614 SAM

```

=====
Sample Name: S98T002614 DUP                      Date: 01/22/1999 00:29:09
Data File  : C:\DX\DATA\99012101.D19
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 19              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      41    3000   5Hz   0.00  10.00      30
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.88		0.000	79	351	1	
2	1.03	fluoride	0.680	71	251	1	4.38
3	1.47	chloride	10.287	687	3092	1	3.03
4	1.75	nitrite	179.295	7230	37384	1	2.14
5	2.86	nitrate	359.600	10118	69794	1	-1.04
6	4.17	phosphate	35.010	229	2804	1	5.22
7	5.33	sulfate	17.266	324	4060	1	6.45
8	6.93	oxalate	38.591	414	6439	1	7.44
Totals			640.729	19151	124174		

File: 99012101.D19 Sample: S98T002614 DUP

```

=====
Sample Name: S98T002614 SPK                      Date: 01/22/1999 00:42:18
Data File  : C:\DX\DATA\99012101.D20
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 20              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.88		0.000	143	631	2	
2	1.03	fluoride	28.192	1818	10566	2	4.38
3	1.49	chloride	42.243	2758	12862	1	3.96
4	1.76	nitrite	398.073	15915	83258	1	2.92
5	2.57	bromide	229.678	5919	32661	1	0.26
6	2.86	nitrate	628.714	17687	124046	1	-1.04
7	4.18	phosphate	255.053	1667	20823	1	5.56
8	5.33	sulfate	273.791	5614	65346	1	6.45
9	6.88	oxalate	268.210	2793	45098	1	6.61
Totals			2123.954	54314	395289		

File: 99012101.D20 Sample: S98T002614 SPK

LABCORE Completed Worklist Report for Worklist# 27733

Analyst: adp

Instrument: IC40S1

Book#: 51N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC	F	QC	1	2.20e-02	0.022 ug/mL
1	CCB	0	@IC-QC	CL	QC	1	<1.70e-2	ug/mL
1	CCB	0	@IC-QC	NO2	QC	1	<1.08e-1	ug/mL
1	CCB	0	@IC-QC	BR	QC	1	<1.25e-1	ug/mL
1	CCB	0	@IC-QC	NO3	QC	1	<1.39e-1	ug/mL
1	CCB	0	@IC-QC	P04	QC	1	<1.20e-1	ug/mL
1	CCB	0	@IC-QC	S04	QC	1	<1.38e-1	ug/mL
1	CCB	0	@IC-QC	OXALATE2	QC	1	<1.05e-1	ug/mL
2	LCS-INST	0	@IC-QC	F	QC	5.52e1	6.41e+01	108.277 % Recovery
2	LCS-INST	0	@IC-QC	CL	QC	7.96e1	8.25e+01	103.643 % Recovery
2	LCS-INST	0	@IC-QC	NO2	QC	5.35e2	4.77e+02	89.159 % Recovery
2	LCS-INST	0	@IC-QC	BR	QC	5.86e2	5.67e+02	96.758 % Recovery
2	LCS-INST	0	@IC-QC	NO3	QC	5.92e2	5.34e+02	107.095 % Recovery
2	LCS-INST	0	@IC-QC	P04	QC	5.49e2	5.59e+02	102.007 % Recovery
2	LCS-INST	0	@IC-QC	S04	QC	6.38e2	6.29e+02	98.589 % Recovery
2	LCS-INST	0	@IC-QC	OXALATE2	QC	5.52e2	5.98e+02	108.333 % Recovery
3	CCV	0	@IC-QC	F	QC	6.40e1	6.23e+01	97.344 % Recovery
3	CCV	0	@IC-QC	CL	QC	8.95e1	9.08e+01	101.453 % Recovery
3	CCV	0	@IC-QC	NO2	QC	5.37e2	5.21e+02	97.020 % Recovery
3	CCV	0	@IC-QC	BR	QC	6.30e2	6.02e+02	95.556 % Recovery
3	CCV	0	@IC-QC	NO3	QC	7.00e2	6.71e+02	95.857 % Recovery
3	CCV	0	@IC-QC	P04	QC	6.35e2	6.39e+02	100.630 % Recovery
3	CCV	0	@IC-QC	S04	QC	7.00e2	6.68e+02	95.429 % Recovery
3	CCV	0	@IC-QC	OXALATE2	QC	5.33e2	5.62e+02	105.441 % Recovery
4	BLNK-PREP	0	@IC-01	F-02	SOLID	1	<1.20e-2	ug/g
4	BLNK-PREP	0	@IC-01	CL-02	SOLID	1	<1.70e-2	ug/g
4	BLNK-PREP	0	@IC-01	NO2-02	SOLID	1	<1.08e-1	ug/g
4	BLNK-PREP	0	@IC-01	BR-02	SOLID	1	<1.25e-1	ug/g
4	BLNK-PREP	0	@IC-01	NO3-02	SOLID	1	<1.39e-1	ug/g
4	BLNK-PREP	0	@IC-01	P04-02	SOLID	1	<1.20e-1	ug/g
4	BLNK-PREP	0	@IC-01	S04-02	SOLID	1	<1.38e-1	ug/g
4	BLNK-PREP	0	@IC-01	OXALATE2	SOLID	1	<1.05e-1	ug/g
5	SAMPLE	S98T002849	0 W	@IC-01	CL-02	SOLID	N/A	1.490e+03
5	SAMPLE	S98T002849	0 W	@IC-01	NO2-02	SOLID	N/A	2.018e+04
5	SAMPLE	S98T002849	0 W	@IC-01	BR-02	SOLID	N/A	2.535e+03
5	SAMPLE	S98T002849	0 W	@IC-01	NO3-02	SOLID	N/A	2.931e+05
5	SAMPLE	S98T002849	0 W	@IC-01	S04-02	SOLID	N/A	3.398e+03
5	SAMPLE	S98T002849	0 W	@IC-01	OXALATE2	SOLID	N/A	4.814e+03
6	DUP	S98T002849	0 W	@IC-01	F-02	SOLID	2.02e+03	6.84e+02
6	DUP	S98T002849	0 W	@IC-01	CL-02	SOLID	1.49e+03	1.39e+03
6	DUP	S98T002849	0 W	@IC-01	NO2-02	SOLID	2.02e+04	1.76e+04

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27733

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002849	0 W	①C-01	BR-02	SOLID	<2.54e3	<2.50e3	RPD
6 DUP	S98T002849	0 W	①C-01	N03-02	SOLID	2.93e+05	3.44e+05	16.013 RPD
6 DUP	S98T002849	0 W	①C-01	P04-02	SOLID	1.87e+04	5.27e+04	112.057 RPD
6 DUP	S98T002849	0 W	①C-01	S04-02	SOLID	3.40e+03	<2.76e3	RPD
6 DUP	S98T002849	0 W	①C-01	OXALATE2	SOLID	4.81e+03	4.00e+03	18.388 RPD
7 SAMPLE	S98T002931	0 W	①C-01	F-02	SOLID	N/A	8.016e+02	240.900 ug/g
7 SAMPLE	S98T002931	0 W	①C-01	N02-02	SOLID	N/A	8.810e+03	2168.000 ug/g
7 SAMPLE	S98T002931	0 W	①C-01	BR-02	SOLID	N/A	2.509e+03	2510.000 ug/g
7 SAMPLE	S98T002931	0 W	①C-01	N03-02	SOLID	N/A	5.367e+05	2791.000 ug/g
8 DUP	S98T002931	0 W	①C-01	F-02	SOLID	8.02e+02	7.98e+02	1.761 RPD
8 DUP	S98T002931	0 W	①C-01	CL-02	SOLID	8.67e+02	6.72e+02	25.341 RPD
8 DUP	S98T002931	0 W	①C-01	N02-02	SOLID	8.81e+03	8.06e+03	8.892 RPD
8 DUP	S98T002931	0 W	①C-01	BR-02	SOLID	<2.51e3	<2.46e3	RPD
8 DUP	S98T002931	0 W	①C-01	N03-02	SOLID	5.37e+05	5.79e+05	7.527 RPD
8 DUP	S98T002931	0 W	①C-01	P04-02	SOLID	4.26e+03	5.79e+03	30.448 RPD
8 DUP	S98T002931	0 W	①C-01	S04-02	SOLID	4.06e+03	3.18e+03	24.309 RPD
8 DUP	S98T002931	0 W	①C-01	OXALATE2	SOLID	4.72e+03	3.41e+03	32.226 RPD
9 SAMPLE	S98T002932	0 W	①C-01	F-02	SOLID	N/A	1.717e+03	242.200 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	CL-02	SOLID	N/A	4.271e+02	343.100 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	N02-02	SOLID	N/A	4.097e+03	2180.000 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	BR-02	SOLID	N/A	2.523e+03	2523.000 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	N03-02	SOLID	N/A	6.549e+05	2806.000 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	P04-02	SOLID	N/A	3.812e+03	2422.000 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	S04-02	SOLID	N/A	9.010e+03	2786.000 ug/g
9 SAMPLE	S98T002932	0 W	①C-01	OXALATE2	SOLID	N/A	3.197e+03	2119.000 ug/g
10 DUP	S98T002932	0 W	①C-01	F-02	SOLID	1.72e+03	1.77e+03	2.865 RPD
10 DUP	S98T002932	0 W	①C-01	CL-02	SOLID	4.27e+02	<3.48e2	RPD
10 DUP	S98T002932	0 W	①C-01	N02-02	SOLID	4.10e+03	4.29e+03	4.529 RPD
10 DUP	S98T002932	0 W	①C-01	BR-02	SOLID	<2.52e3	<2.56e3	RPD
10 DUP	S98T002932	0 W	①C-01	N03-02	SOLID	6.55e+05	6.56e+05	0.153 RPD
10 DUP	S98T002932	0 W	①C-01	P04-02	SOLID	3.81e+03	<2.46e3	RPD
10 DUP	S98T002932	0 W	①C-01	S04-02	SOLID	9.01e+03	9.45e+03	4.767 RPD
10 DUP	S98T002932	0 W	①C-01	OXALATE2	SOLID	3.14e+03	3.32e+03	5.573 RPD
11 SPK	S98T002932	0 W	①C-01	F-02	SOLID	5.92e1	6.36e+01	107.432 % Recovery
11 SPK	S98T002932	0 W	①C-01	CL-02	SOLID	7.96e1	7.56e+01	94.975 % Recovery
11 SPK	S98T002932	0 W	①C-01	N02-02	SOLID	5.35e2	4.55e+02	85.047 % Recovery
11 SPK	S98T002932	0 W	①C-01	BR-02	SOLID	5.86e2	5.46e+02	93.174 % Recovery
11 SPK	S98T002932	0 W	①C-01	N03-02	SOLID	5.92e2	6.50e+02	109.797 % Recovery
11 SPK	S98T002932	0 W	①C-01	P04-02	SOLID	5.48e2	5.44e+02	99.270 % Recovery
11 SPK	S98T002932	0 W	①C-01	S04-02	SOLID	6.38e2	6.17e+02	96.708 % Recovery
11 SPK	S98T002932	0 W	①C-01	OXALATE2	SOLID	5.52e2	5.83e+02	105.616 % Recovery

Final page for worklist# 27733

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jann M. Lye 2/9/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

12/30/98 14:52
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27733

Analyst: ADP Instrument: IC 4051 Book# 51N21-C LCS
 Method: LA-533-105 Rev/Mod F-6 50N21-C CCV
 Worklist Comment: BY-105 FOR @IC-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002849 0 W	@IC-01	SOLID		98000437	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
6 DUP	S98T002849 0 W	@IC-01	SOLID			
7 SAMPLE	S98T002931 0 W	@IC-01	SOLID		98000490	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
8 DUP	S98T002931 0 W	@IC-01	SOLID			
9 SAMPLE	S98T002932 0 W	@IC-01	SOLID		98000490	BY-105 (2)
Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
10 DUP	S98T002932 0 W	@IC-01	SOLID			
11 SPK	S98T002932 0 W	@IC-01	SOLID			

Final page for worklist # 27733

Anthony Purnh 2-7-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

Review mis-data uploaded on 2/8/99
S98T002849 & S98T002931 Reported for some
analyses. These will be re-run because of
high RPD's. Rest validated, M Lye 2/8/99
Reanalyzed on wk# 28270. J Star 3/24/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: **50N21-C LCS** Date: 02/07/1999 19:36:44
 Data File : C:\DX\DATA\99020701.D02
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 2 Detector: CDM-1
 Analyst : *Anthony Parvate* Column: AG4A/AS4A anion column

02-07-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	78	285	2	
2	0.99	fluoride	64.129	108.33	1935	9743	-0.34
3	1.43	chloride	82.470	103.61	2270	10157	-0.23
4	1.69	nitrite	476.924	89.15	8064	40376	-1.36
5	2.45	bromide	567.095	96.77	6024	32737	-4.17
6	2.76	nitrate	634.411	107.16	7831	49684	0.00
7	4.04	phosphate	559.333	102.07	1532	18491	2.02
8	5.23	sulfate	629.074	98.60	5168	60881	4.32
9	6.77	oxalate	598.045	108.34	2592	40785	4.96
Totals			3611.480	35493	263138		

File: 99020701.D02 Sample: **50N21-C LCS**
51
1400 2-7-99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1481 TO 1491.

Sample Name: 50N21-C ~~LCS~~ ~~CCV~~ Date: 02/07/1999 19:49:11
 Data File : C:\DX\DATA\99020701.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	82	312	2	
2	0.99	fluoride	62.252	97.27	1891	9454	2 -0.34
3	1.43	chloride	90.753	101.4	2501	11192	1 -0.23
4	1.69	nitrite	520.964	97.01	8833	44115	1 -1.36
5	2.45	bromide	601.667	95.50	6370	34767	1 -4.43
6	2.75	nitrate	670.702	95.81	8259	52571	1 -0.24
7	4.04	phosphate	639.047	100.64	1775	21188	1 2.02
8	5.23	sulfate	668.284	95.47	5561	64738	1 4.32
9	6.77	oxalate	562.393	165.51	2444	38335	1 4.96
Totals			3816.062	37716	276671		

File: 99020701.D03 Sample: 50N21-C LCS

```

=====
Sample Name: BLANK                      Date: 02/07/1999 19:24:36
Data File  : C:\DX\DATA\99020701.D01
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 1      Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.02	fluoride	0.022	53	327	1	3.03
Totals			0.022	53	327		

File: 99020701.D01 Sample: BLANK

```
=====
Sample Name: PREP BLANK                      Date: 02/07/1999 20:00:52
Data File  : C:\DX\DATA\99020701.D04
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 4           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====
```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.26		0.000	54	462	1	
2	1.44	chloride	0.013	43	158	1	0.70
3	2.80	nitrate	0.009	13	72	1	1.45
Totals			0.022	110	692		

File: 99020701.D04 Sample: PREP BLANK

Sample Name: S98T002849	SAM	Date: 02/07/1999 22:15:22
Data File :	F:\DATA\99020701.D09	
Method :	C:\DX\METHOD\KIT.MET	
ACI Address: 1	System: 1	Inject#: 9
Analyst :	Column: AG4A/AS4A anion column	Detector: CDM-1

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	60	244	2	
2	0.99	fluoride	10.053	288	1507	2	-0.34
3	1.43	chloride	7.421	215	903	1	-0.23
4	1.68	nitrite	100.540	1705	8493	1	-1.75
5	2.72	nitrate	1459.999	17322	116672	1	0.00
6	4.04	phosphate	93.167	250	3029	1	2.02
7	5.23	sulfate	16.924	131	1614	1	4.32
8	6.77	oxalate	23.978	103	1622	1	4.96
Totals			1712.083	20075	134087		

Data Reprocessed On 02/08/1999 13:16:41

```

=====
Sample Name: S98T002849  DUP                      Date: 02/07/1999 22:26:51
Data File  : F:\DATA\99020701.D10
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 10             Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101      3000  5Hz  0.00 10.00      30
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	62	262	2	
2	0.99	fluoride	3.448	113	516	2	0.34
3	1.43	chloride	7.007	198	852	1	-0.23
4	1.69	nitrite	88.581	1520	7483	1	-1.36
5	2.72	nitrate	1735.655	20282	139680	1	0.00
6	4.06	phosphate	26.568	87	862	1	2.53
7	5.23	sulfate	13.039	88	1244	1	4.32
8	6.77	oxalate	20.188	77	1366	1	4.96
Totals			1894.486	22426	152265		

File: 99020701.D10 Sample: S98T002849 DUP

```

=====
Sample Name: S98T002931 SAM                      Date: 02/07/1999 22:38:50
Data File  : F:\DATA\99020701.D11
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 11              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	69	273	2	
2	0.99	fluoride	4.033	140	604	2	-0.34
3	1.43	chloride	4.361	125	530	1	0.23
4	1.68	nitrite	44.322	719	3743	1	-1.75
5	2.70	nitrate	2700.176	31217	223002	1	0.00
6	4.06	phosphate	21.446	67	696	1	2.53
7	5.23	sulfate	20.449	122	1951	1	4.32
8	6.77	oxalate	23.764	82	1608	1	4.96
Totals			2818.551	32540	232406		

File: 99020701.D11 Sample: S98T002931 SAM

Data Reprocessed On 02/08/1999 13:18:17

```

=====
Sample Name: S98T002931  DUP                      Date: 02/07/1999 22:50:29
Data File  : F:\DATA\99020701.D12
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 12             Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	69	268	2	
2	0.99	fluoride	4.053	150	607	2	-0.34
3	1.43	chloride	3.453	97	420	1	-0.23
4	1.68	nitrite	41.443	693	3500	1	-1.75
5	2.69	nitrate	2975.296	34163	247644	1	0.12
6	4.05	phosphate	29.758	89	966	1	2.19
7	5.23	sulfate	16.327	104	1557	1	4.32
8	6.77	oxalate	17.555	71	1188	1	4.96
Totals			3087.884	35435	256149		

File: 99020701.D12 Sample: S98T002931 DUP

```

=====
Sample Name: S98T002932 SAM                      Date: 02/07/1999 23:01:27
Data File  : F:\DATA\99020701.D13
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 13              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	77	283	2	
2	0.99	fluoride	8.594	289	1288	2	-0.34
3	1.43	chloride	2.137	61	260	1	0.23
4	1.68	nitrite	20.502	353	1731	1	-1.75
5	2.69	nitrate	3277.334	37303	275189	1	-0.12
6	4.05	phosphate	19.076	61	619	1	2.36
7	5.23	sulfate	45.085	330	4303	1	4.32
8	6.77	oxalate	15.696	60	1062	1	4.96
Totals			3388.425	38533	284736		

File: 99020701.D13 Sample: S98T002932 SAM

Data Reprocessed On 02/08/1999 13:19:42

```

=====
Sample Name: S98T002932 DUP                      Date: 02/07/1999 23:12:39
Data File  : F:\DATA\99020701.D14
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 14              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	69	262	2	
2	0.99	fluoride	8.757	289	1313	2	-0.34
3	1.43	chloride	1.625	50	197	1	-0.23
4	1.69	nitrite	21.175	371	1788	1	-1.36
5	2.69	nitrate	3235.481	36382	271341	1	0.12
6	4.07	phosphate	11.972	43	388	1	2.69
7	5.23	sulfate	46.660	335	4454	1	4.32
8	6.77	oxalate	16.404	59	1110	1	4.96
Totals			3342.074	37598	280853		

File: 99020701.D14 Sample: S98T002932 DUP

Sample Name: S98T002932 SPK Date: 02/07/1999 23:25:26
 Data File : F:\DATA\99020701.D15
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 15 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	135	481	2	
2	0.99	fluoride	72.207	2218	10993	2	-0.34
3	1.43	chloride	77.781	2196	9572	1	-0.23
4	1.68	nitrite	475.157	8102	40226	1	-1.75
5	2.43	bromide	545.530	6087	31473	1	0.14
6	2.67	nitrate	3927.133	44525	336334	1	0.12
7	4.03	phosphate	563.115	1557	18618	1	1.85
8	5.23	sulfate	661.610	5515	64081	1	4.32
9	6.72	oxalate	599.143	2384	40860	1	4.13
Totals			6921.677	72720	552637		

File: 99020701.D15 Sample: S98T002932 SPK

LABCORE Completed Worklist Report for Worklist# 27734

Analyst: adp

Instrument: IC40S1

Book#: 61N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	②IC-QC	F	QC	1	<1.20e-2	ug/mL
1	CCB	0	②IC-QC	CL	QC	1	<1.70e-2	ug/mL
1	CCB	0	②IC-QC	NO2	QC	1	<1.08e-1	ug/mL
1	CCB	0	②IC-QC	BR	QC	1	<1.25e-1	ug/mL
1	CCB	0	②IC-QC	NO3	QC	1	<1.39e-1	ug/mL
1	CCB	0	②IC-QC	P04	QC	1	<1.20e-1	ug/mL
1	CCB	0	②IC-QC	S04	QC	1	<1.38e-1	ug/mL
1	CCB	0	②IC-QC	OXALATE2	QC	1	<1.05e-1	ug/mL
2	LCS-INST	0	②IC-QC	F	QC	5.920e1	6.46e+01	109.122 % Recovery
2	LCS-INST	0	②IC-QC	CL	QC	7.960e1	7.87e+01	98.869 % Recovery
2	LCS-INST	0	②IC-QC	NO2	QC	5.480e2	4.97e+02	90.693 % Recovery
2	LCS-INST	0	②IC-QC	BR	QC	5.860e2	5.77e+02	98.464 % Recovery
2	LCS-INST	0	②IC-QC	NO3	QC	5.920e2	6.01e+02	101.520 % Recovery
2	LCS-INST	0	②IC-QC	P04	QC	5.480e2	5.31e+02	96.898 % Recovery
2	LCS-INST	0	②IC-QC	S04	QC	6.280e2	6.27e+02	98.276 % Recovery
2	LCS-INST	0	②IC-QC	OXALATE2	QC	5.530e2	6.04e+02	109.222 % Recovery
3	CCV	0	②IC-QC	F	QC	6.40e1	6.20e+01	96.875 % Recovery
3	CCV	0	②IC-QC	CL	QC	8.95e1	8.78e+01	98.101 % Recovery
3	CCV	0	②IC-QC	NO2	QC	5.54e2	5.58e+02	100.722 % Recovery
3	CCV	0	②IC-QC	BR	QC	6.30e2	6.09e+02	96.667 % Recovery
3	CCV	0	②IC-QC	NO3	QC	7.00e2	6.73e+02	96.143 % Recovery
3	CCV	0	②IC-QC	P04	QC	6.35e2	6.26e+02	98.583 % Recovery
3	CCV	0	②IC-QC	S04	QC	7.00e2	6.75e+02	96.429 % Recovery
3	CCV	0	②IC-QC	OXALATE2	QC	5.49e2	5.95e+02	108.379 % Recovery
4	BLNK-PREP	0	②IC-01	F-02	SOLID	1	<1.20e-2	ug/g
4	BLNK-PREP	0	②IC-01	CL-02	SOLID	1	<1.70e-2	ug/g
4	BLNK-PREP	0	②IC-01	NO2-02	SOLID	1	<1.08e-1	ug/g
4	BLNK-PREP	0	②IC-01	BR-02	SOLID	1	<1.25e-1	ug/g
4	BLNK-PREP	0	②IC-01	NO3-02	SOLID	1	<1.39e-1	ug/g
4	BLNK-PREP	0	②IC-01	P04-02	SOLID	1	<1.20e-1	ug/g
4	BLNK-PREP	0	②IC-01	S04-02	SOLID	1	<1.38e-1	ug/g
4	BLNK-PREP	0	②IC-01	OXALATE2	SOLID	1	<1.05e-1	ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	CL-02	SOLID	N/A	3.989e+02 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	NO2-02	SOLID	N/A	2.710e+03 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	BR-02	SOLID	N/A	2.502e+03 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	NO3-02	SOLID	N/A	7.182e+05 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	P04-02	SOLID	N/A	2.401e+03 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	S04-02	SOLID	N/A	3.614e+03 ug/g
5	SAMPLE	S98T002933	0 W	②IC-01	OXALATE2	SOLID	N/A	2.101e+03 ug/g
6	DUP	S98T002933	0 W	②IC-01	F-02	SOLID	4.25e+02	6.53e+02 42.301 RPD
6	DUP	S98T002933	0 W	②IC-01	CL-02	SOLID	3.99e+02	<3.41e2 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27734

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002933	0 W	QIC-01	NO2-02	SOLID	2.71e+03	3.01e+03	10.490 RPD
6 DUP	S98T002933	0 W	QIC-01	BR-02	SOLID	<2.50e3	<2.51e3	RPD
6 DUP	S98T002933	0 W	QIC-01	NO3-02	SOLID	7.18e+05	7.09e+05	1.261 RPD
6 DUP	S98T002933	0 W	QIC-01	PO4-02	SOLID	<2.40e3	<2.41e3	RPD
6 DUP	S98T002933	0 W	QIC-01	SO4-02	SOLID	3.61e+03	3.50e+03	0.277 RPD
6 DUP	S98T002933	0 W	QIC-01	OXALATE2	SOLID	<2.10e3	<2.11e3	RPD
7 SPK	S98T002933	0 W	QIC-01	F-02	SOLID	5.920e1	5.87e+01	99.155 % Recovery
7 SPK	S98T002933	0 W	QIC-01	CL-02	SOLID	7.960e1	7.46e+01	93.719 % Recovery
7 SPK	S98T002933	0 W	QIC-01	NO2-02	SOLID	5.480e2	4.80e+02	87.591 % Recovery
7 SPK	S98T002933	0 W	QIC-01	BR-02	SOLID	5.860e2	5.45e+02	93.003 % Recovery
7 SPK	S98T002933	0 W	QIC-01	NO3-02	SOLID	5.920e2	5.27e+02	89.020 % Recovery
7 SPK	S98T002933	0 W	QIC-01	PO4-02	SOLID	5.480e2	5.22e+02	95.255 % Recovery
7 SPK	S98T002933	0 W	QIC-01	SO4-02	SOLID	6.380e2	5.98e+02	93.730 % Recovery
7 SPK	S98T002933	0 W	QIC-01	OXALATE2	SOLID	5.530e2	5.83e+02	105.425 % Recovery
8 SAMPLE	S98T002934	0 W	QIC-01	CL-02	SOLID	N/A	9.125e+02	335.500 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	NO2-02	SOLID	N/A	1.204e+04	2131.000 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	BR-02	SOLID	N/A	2.467e+03	2466.000 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	NO3-02	SOLID	N/A	6.259e+05	2743.000 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	PO4-02	SOLID	N/A	2.368e+03	2368.000 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	SO4-02	SOLID	N/A	3.227e+03	2723.000 ug/g
8 SAMPLE	S98T002934	0 W	QIC-01	OXALATE2	SOLID	N/A	2.072e+03	2073.000 ug/g
9 DUP	S98T002934	0 W	QIC-01	F-02	SOLID	5.19e+02	6.58e+02	23.619 RPD
9 DUP	S98T002934	0 W	QIC-01	CL-02	SOLID	9.12e+02	9.02e+02	1.103 RPD
9 DUP	S98T002934	0 W	QIC-01	NO2-02	SOLID	1.20e+04	1.07e+04	11.454 RPD
9 DUP	S98T002934	0 W	QIC-01	BR-02	SOLID	<2.47e3	<2.48e3	RPD
9 DUP	S98T002934	0 W	QIC-01	NO3-02	SOLID	6.26e+05	6.20e+05	0.963 RPD
9 DUP	S98T002934	0 W	QIC-01	PO4-02	SOLID	<2.37e3	<2.38e3	RPD
9 DUP	S98T002934	0 W	QIC-01	SO4-02	SOLID	3.23e+03	3.07e+03	5.079 RPD
9 DUP	S98T002934	0 W	QIC-01	OXALATE2	SOLID	<2.07e3	<2.09e3	RPD
11 DUP	S98T002935	0 W	QIC-01	F-02	SOLID	4.65e+02	7.67e+02	49.026 RPD
11 DUP	S98T002935	0 W	QIC-01	CL-02	SOLID	1.07e+03	1.35e+03	23.140 RPD
11 DUP	S98T002935	0 W	QIC-01	NO2-02	SOLID	1.08e+04	1.38e+04	24.390 RPD
11 DUP	S98T002935	0 W	QIC-01	BR-02	SOLID	<2.54e3	<2.54e3	RPD
11 DUP	S98T002935	0 W	QIC-01	NO3-02	SOLID	4.91e+05	3.46e+05	34.648 RPD
11 DUP	S98T002935	0 W	QIC-01	PO4-02	SOLID	<2.44e3	<2.44e3	RPD
11 DUP	S98T002935	0 W	QIC-01	SO4-02	SOLID	6.05e+03	8.56e+03	34.360 RPD
11 DUP	S98T002935	0 W	QIC-01	OXALATE2	SOLID	7.93e+03	1.34e+04	51.289 RPD
12 SPK	S98T002935	0 W	QIC-01	F-02	SOLID	5.920e1	6.05e+01	102.196 % Recovery
12 SPK	S98T002935	0 W	QIC-01	CL-02	SOLID	7.960e1	7.64e+01	95.980 % Recovery
12 SPK	S98T002935	0 W	QIC-01	NO2-02	SOLID	5.480e2	5.05e+02	92.153 % Recovery
12 SPK	S98T002935	0 W	QIC-01	BR-02	SOLID	5.860e2	5.38e+02	91.809 % Recovery
12 SPK	S98T002935	0 W	QIC-01	NO3-02	SOLID	5.920e2	5.31e+02	89.696 % Recovery
12 SPK	S98T002935	0 W	QIC-01	PO4-02	SOLID	5.480e2	5.23e+02	95.438 % Recovery
12 SPK	S98T002935	0 W	QIC-01	SO4-02	SOLID	6.380e2	6.11e+02	95.768 % Recovery
12 SPK	S98T002935	0 W	QIC-01	OXALATE2	SOLID	5.530e2	5.92e+02	107.052 % Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27734

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 27734

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John L. Lyle 2/25/99
Reviewer Signature Date

12/30/98 14:52
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27734

Analyst: ADP Instrument: IC 4051 Book# 60N21-A ECS-CCV
 Method: LA-533-105 Rev/Mod F-D 61N21-B LCS
 Worklist Comment: BY-105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T002933	0	W	@IC-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
6	DUP	S98T002933	0	W	@IC-01	SOLID		
7	SPK	S98T002933	0	W	@IC-01	SOLID		
8	SAMPLE	S98T002934	0	W	@IC-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
9	DUP	S98T002934	0	W	@IC-01	SOLID		
10	SAMPLE	S98T002935	0	W	@IC-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
11	DUP	S98T002935	0	W	@IC-01	SOLID		
12	SPK	S98T002935	0	W	@IC-01	SOLID		

Final page for worklist # 27734

Anthony Furman 2-14-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 2-18-99

Dr. Warrall

27734 FEB.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Rejected all of 598T002935 and F

only me 598T002933 & 598T002934

low high RPD's. Rest validated

GM Faye 2/25/99

598T002933 + 2934 Reanalyzed on WL# 28394. Later 3/24/99

Sample Name: 61N21-B LCS Date: 02/14/1999 08:32:22
 Data File : C:\DX\DATA\99021401.D31
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 31 Detector: CDM-1
 Analyst: *Arthur Furness* Column: AG4A/AS4A anion column

2-14-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	88	324	2	
2	0.99	fluoride	64.559	1903	9810	2	-0.34
3	1.42	chloride	78.655	2273	9681	1	-0.70
4	1.67	nitrite	496.651	8542	42051	1	-2.14
5	2.43	bromide	577.362	5973	33340	1	-5.21
6	2.73	nitrate	600.921	7216	47024	1	-0.97
7	4.00	phosphate	531.131	1435	17540	1	1.01
8	5.12	sulfate	627.399	5356	60717	1	2.20
9	6.61	oxalate	604.143	2718	41204	1	2.48
Totals			3580.821	35505	261689		

File: 99021401.D31 Sample: 61N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1496 TO 1504.

```

=====
Sample Name: 60N21-A   CCV                               Date: 02/14/1999 08:44:51
Data File  : C:\DX\DATA\99021401.D32
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 32                      Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000 5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	85	292	2	
2	0.99	fluoride	62.040	1865	9421	2	-0.34
3	1.42	chloride	87.750	2538	10816	1	-0.70
4	1.68	nitrite	558.496	9442	47304	1	-1.75
5	2.43	bromide	609.374	6504	35220	1	-5.21
6	2.73	nitrate	673.138	8286	52765	1	-0.97
7	4.00	phosphate	626.343	1698	20757	1	1.01
8	5.07	sulfate	674.583	4504	65358	1	1.13
9	6.61	oxalate	595.440	2648	40606	1	2.48
Totals			3887.164	37572	282538		

File: 99021401.D32 Sample: 60N21-A CCV

```

=====
Sample Name: BLANK                               Date: 02/14/1999 08:20:39
Data File  : C:\DX\DATA\99021401.D30
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 30             Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	2.85	nitrate	0.059	53	457	1	3.38
Totals			0.059	53	457		

File: 99021401.D30 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 02/14/1999 08:56:06
Data File  : C:\DX\DATA\99021401.D33
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 33          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.21		0.000	21	75	1	
2	1.47	chloride	0.004	20	49	1	2.56
3	2.81	nitrate	0.010	12	76	1	1.69
Totals			0.014	53	200		

File: 99021401.D33 Sample: PREP BLANK

Data Reprocessed On 02/18/1999 12:41:08

```

=====
Sample Name: S98T002933  SAM                      Date: 02/14/1999 09:20:34
Data File  : F:\DATA\99021401.D35
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 35              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	85	326	2	
2	0.99	fluoride	2.144	88	321	2	-0.34
3	1.42	chloride	2.013	45	245	1	-0.70
4	1.67	nitrite	13.677	248	1155	1	-2.14
5	2.66	nitrate	3624.917	41337	307562	1	0.00
6	5.12	sulfate	18.238	146	1740	1	2.20
7	6.67	oxalate	6.427	23	435	1	3.31
Totals			3667.417	41972	311783		

File: 99021401.D35 Sample: S98T002933 SAM

Data Reprocessed On 02/18/1999 12:41:49

```

=====
Sample Name: S98T002933  DUP                      Date: 02/14/1999 09:33:37
Data File  : F:\DATA\99021401.D36
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 36                Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	94	420	2	
2	0.99	fluoride	3.289	105	492	2	-0.34
3	1.42	chloride	1.244	36	151	1	-0.70
4	1.67	nitrite	15.147	269	1279	1	-2.14
5	2.67	nitrate	3566.059	40723	302027	1	0.00
6	5.12	sulfate	18.100	136	1727	1	2.20
7	6.61	oxalate	7.500	29	507	1	2.48
Totals			3611.340	41392	306604		

File: 99021401.D36 Sample: S98T002933 DUP

Data Reprocessed On 02/24/1999 10:22:11

```

=====
Sample Name: S98T002933  SPK                      Date: 02/14/1999 09:45:10
Data File  : F:\DATA\99021401.D37
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 37             Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	155	573	2	
2	0.99	fluoride	60.854	1918	9238	2	-0.34
3	1.42	chloride	76.606	2197	9425	1	-0.70
4	1.67	nitrite	493.280	8505	41764	1	-2.14
5	2.41	bromide	545.269	6002	31458	1	0.14
6	2.65	nitrate	4151.775	47019	358120	1	0.00
7	4.00	phosphate	521.832	1403	17227	1	1.01
8	5.07	sulfate	615.978	4271	59595	1	1.13
9	6.56	oxalate	583.411	2385	39779	1	1.65
Totals			7049.006	73856	567181		

File: 99021401.D37 Sample: S98T002933 SPK

Data Reprocessed On 02/18/1999 12:43:18

```

=====
Sample Name: S98T002934 SAM                      Date: 02/14/1999 10:19:42
Data File  : F:\DATA\99021401.D39
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 39              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	92	393	2	
2	0.99	fluoride	2.654	100	397	2	-0.34
3	1.42	chloride	4.670	133	568	1	-0.70
4	1.67	nitrite	61.601	1044	5203	1	-2.14
5	2.67	nitrate	3203.474	36446	268405	1	0.00
6	4.01	phosphate	6.632	28	215	1	1.18
7	5.12	sulfate	16.518	131	1576	1	2.20
Totals			3295.549	37974	276756		

File: 99021401.D39 Sample: S98T002934 SAM

Data Reprocessed On 02/18/1999 12:43:58

```

=====
Sample Name: S98T002934  DUP                      Date: 02/14/1999 10:30:22
Data File  : F:\DATA\99021401.D40
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 40              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	86	392	2	
2	0.99	fluoride	3.348	100	501	2	-0.34
3	1.43	chloride	4.588	126	558	1	-0.23
4	1.68	nitrite	54.643	932	4615	1	-1.75
5	2.67	nitrate	3152.816	35606	263770	1	0.00
6	4.02	phosphate	4.464	20	145	1	1.52
7	5.12	sulfate	15.619	117	1490	1	2.20
Totals			3235.477	36987	271470		

File: 99021401.D40 Sample: S98T002934 DUP

LABCORE Completed Worklist Report for Worklist# 27735

Analyst: vlm

Instrument: IC45S2

Book#: 35N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.05e1	6.46e+01	106.777 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.19e2	9.47e+01	79.580 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.28e2	1.21e+02	94.531 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLTI	QC	1.16e2	1.15e+02	99.138 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.21e1	5.99e+01	96.457 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.17e+02	95.122 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.23e+02	94.615 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLTI	QC	1.02e2	9.93e+01	97.353 % Recovery	
4 BLNK-PREP	0	@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002849	0 W @IC4G-01 F*4-01	SOLID	N/A	2.204e+03	130.700	ug/g
5 SAMPLE	S98T002849	0 W @IC4G-01 GLYCOLTI	SOLID	N/A	< 2.361e+02	236.000	ug/g
6 DUP	S98T002849	0 W @IC4G-01 F*4-01	SOLID	2.20e+03	2.22e+03	0.905	RFD
6 DUP	S98T002849	0 W @IC4G-01 ACETATE2	SOLID	6.54e+02	8.65e+01	153.275	RFD
6 DUP	S98T002849	0 W @IC4G-01 FORMATE2	SOLID	2.66e+02	<2.59e2		RFD
6 DUP	S98T002849	0 W @IC4G-01 GLYCOLTI	SOLID	<2.36e2	<2.33e2		RFD
7 SAMPLE	S98T002932	0 W @IC4G-01 ACETATE2	SOLID	N/A	< 8.333e+01	83.940	ug/g
7 SAMPLE	S98T002932	0 W @IC4G-01 FORMATE2	SOLID	N/A	< 2.602e+02	260.200	ug/g
7 SAMPLE	S98T002932	0 W @IC4G-01 GLYCOLTI	SOLID	N/A	< 2.350e+02	235.000	ug/g
8 DUP	S98T002932	0 W @IC4G-01 F*4-01	SOLID	2.10e+03	2.27e+03	7.780	RFD
8 DUP	S98T002932	0 W @IC4G-01 ACETATE2	SOLID	<8.39e1	<8.51e1		RFD
8 DUP	S98T002932	0 W @IC4G-01 FORMATE2	SOLID	<2.60e2	<2.64e2		RFD
8 DUP	S98T002932	0 W @IC4G-01 GLYCOLTI	SOLID	<2.35e2	<2.38e2		RFD
9 SPK	S98T002932	0 W @IC4G-01 F*4-01	SOLID	6.05e1	7.64e+01	126.281 % Recovery	
9 SPK	S98T002932	0 W @IC4G-01 ACETATE2	SOLID	1.19e2	1.00e+02	84.034 % Recovery	
9 SPK	S98T002932	0 W @IC4G-01 FORMATE2	SOLID	1.28e2	1.24e+02	96.875 % Recovery	
9 SPK	S98T002932	0 W @IC4G-01 GLYCOLTI	SOLID	1.16e2	1.14e+02	98.276 % Recovery	

Final page for worklist# 27735

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Janis M. Lye 1/25/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27735

Analyst: VLM Instrument: IC 4532

Book# LCS 35N21A
CEV 34N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002849 0 W		@IC4G-01	SOLID	98000437	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
6 DUP	S98T002849 0 W		@IC4G-01	SOLID		
7 SAMPLE	S98T002932 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
8 DUP	S98T002932 0 W		@IC4G-01	SOLID		
9 SPK	S98T002932 0 W		@IC4G-01	SOLID		

Final page for worklist # 27735

See Attached

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Data Entry Comments:

1-25-99
John Wall
27735JAN.CSV

Portions of each reported for RPD
and spk problems. JM Lugo

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1506

Reanalyzed on W/L # 28346.
1/25/99
3/24/99

LABCORE Data Entry Template for Worklist# 27735

Analyst: YLM Instrument: IC 4552

Book# LCS 35N21-A
CCV 34N21-A

Method: LA-533-105 Rev/Mod EO

Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002849	0 W	@IC4G-01	SOLID	98000437	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
6 DUP	S98T002849	0 W	@IC4G-01	SOLID		
7 SAMPLE	S98T002931	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
8 DUP	S98T002931	0 W	@IC4G-01	SOLID		
9 SAMPLE	S98T002932	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
10 DUP	S98T002932	0 W	@IC4G-01	SOLID		
11 SPK	S98T002932	0 W	@IC4G-01	SOLID		

Final page for worklist # 27735

Valerie Massie 01-22-99
Analyst Signature Date

Analyst Signature Date

need to make new worklist for sample S98T002931.
Ran out of time!

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: LCS 35N21-A Date: 01/22/1999 11:35:03
 Data File : C:\DX\DATA\99012201.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
 Analyst : *Valerie Massie* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5067 5Hz 0.00 16.89 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
29	7.25	fluoride	106.75	64.583	1826774	27763159	2 0.00
30	7.97	glycolate	93.77	114.576	549099	10208970	2 0.00
31	8.45	acetate	79.55	94.664	493449	9520213	2 0.00
32	9.17	formate	94.80	121.350	1038574	20400387	2 0.00
Totals			395.172	3907895	67892729		

File: 99012201.D03 Sample: LCS 35N21-A

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1508 TO 1521.

Sample Name: CCV 34N21-A Date: 01/22/1999 11:52:20
 Data File : C:\DK\DATA\99012201.D04
 Method : C:\DK\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-2
 Analyst : *Valerie M. Moore* Column: AG14A-SC, AS14A-SC, SRS
 01-22-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4797	5Hz	0.00	15.99	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	7224	185124	1	
29	7.23	fluoride	96.40	59.865	1696882	2	0.00
30	7.95	glycolate	97.34	99.285	475129	2	0.00
31	8.43	acetate	95.46	117.418	597027	2	0.00
32	9.17	formate	94.68	123.081	1062790	2	0.00
Totals			399.648	3839052	67143158		

File: 99012201.D04 Sample: CCV 34N21-A

```

=====
Sample Name: CCB                      Date: 01/22/1999 11:14:50
Data File  : C:\DX\DATA\99012201.D02
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 2   Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1  5700  5Hz  0.00 19.00      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
7	3.33		0.000	6120	87121	2	
12	4.01		0.000	6176	66463	2	
13	4.17		0.000	5331	51301	2	
15	4.43		0.000	6272	65623	2	
19	6.05		0.000	1228	79668	1	
20	9.36		0.000	28883	8870781	3	
21	9.97		0.000	1812	419171	4	
Totals			0.000	55822	9640127		

File: 99012201.D02 Sample: CCB

```

=====
Sample Name: PREP BLANK                      Date: 01/22/1999 12:15:13
Data File  : C:\DX\DATA\99012201.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           1 5700 5Hz 0.00 19.00      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
32	7.09	fluoride	0.010	3610	96041	2	0.00
34	8.13		0.000	3069	104037	2	
35	9.20	formate	0.009	3923	107720	2	0.00
Totals			0.020	10602	307797		

File: 99012201.D05 Sample: PREP BLANK

```

=====
Sample Name: S98T002849      Date: 01/22/1999 14:31:20
Data File  : C:\DX\DATA\99012201.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9
Analyst    :                  Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21  5299  5Hz  0.00 17.66   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	2.24		0.000	4096	76304	1	
5	2.96		0.000	4924	95541	2	
21	7.20	fluoride	10.980	1495064	22565025	2	0.00
22	8.37	acetate	3.260	84101	1766836	2	0.00
23	9.12	formate	1.323	52174	1008417	2	0.00
24	9.89		0.000	5534	80426	1	
Totals			15.562	1645893	25592549		

File: 99012201.D09 Sample: S98T002849

Data Reprocessed On 01/25/1999 10:36:50

```

=====
Sample Name: S98T002849                      Date: 01/22/1999 14:31:20
Data File  : E:\DATA\99012201.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====
    
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5299   5Hz   6.71   9.94      50000
    
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	2.24		0.000	4096	76304	1	
5	2.96		0.000	4924	95541	2	
21	7.20	fluoride	10.980	1495064	22565025	2	0.00
22	8.37	acetate	3.260	84101	1766836	2	0.00
23	9.12	formate	1.323	52174	1008417	2	0.00
24	9.89		0.000	5534	80426	1	
Totals			15.562	1645893	25592549		

File: 99012201.D09 Sample: S98T002849

```

=====
Sample Name: S98T002849 DUP                      Date: 01/22/1999 14:49:27
Data File  : C:\DX\DATA\99012201.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 4797 5Hz 0.00 15.99 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	13875	127658	1	
5	2.93		0.000	10063	351084	2	
7	3.43		0.000	6868	57784	2	
8	3.68		0.000	7599	74678	2	
9	3.76		0.000	7797	112426	2	
12	4.43		0.000	4614	76455	2	
13	4.81		0.000	6549	102311	2	
14	5.07		0.000	6479	114755	2	
16	5.49		0.000	4983	50774	2	
17	5.77		0.000	6766	82226	2	
18	5.87		0.000	6800	118969	2	
20	7.17	fluoride	11.185	1508221	22999810	2	0.00
21	8.35	acetate	0.436	19390	429514	2	0.00
22	9.09	formate	0.551	18129	391655	2	0.00

Totals 12.171 1628133 25090098

File: 99012201.D10 Sample: S98T002849 DUP

Data Reprocessed On 01/25/1999 10:37:48

```

=====
Sample Name: S98T002849 DUP                      Date: 01/22/1999 14:49:27
Data File  : E:\DATA\99012201.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
External      1      21    4797 5Hz    6.11 10.39    50000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	13875	127658	1	
5	2.93		0.000	10063	351084	2	
7	3.43		0.000	6868	57784	2	
8	3.68		0.000	7599	74678	2	
9	3.76		0.000	7797	112426	2	
12	4.43		0.000	4614	76455	2	
13	4.81		0.000	6549	102311	2	
14	5.07		0.000	6479	114755	2	
16	5.49		0.000	4983	50774	2	
17	5.77		0.000	6766	82226	2	
18	5.87		0.000	6800	118969	2	
20	7.17	fluoride	11.185	1508221	22999810	2	0.00
21	8.35	acetate	0.436	19390	429514	2	0.00
22	9.09	formate	0.551	18129	391655	2	0.00
Totals			12.171	1628133	25090098		

File: 99012201.D10 Sample: S98T002849 DUP

Sample Name: S98T002932 Date: 01/22/1999 13:34:14
 Data File : C:\DX\DATA\99012201.D06
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 6 Detector: CDM-2
 Analyst : *Valerie M. Masini* Column: AG14A-SC, AS14A-SC, SRS
01-22-99

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 21 5017 5Hz 0.00 16.72 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.70		0.000	3787	86535	2	
3	1.70		0.000	19901	202212	1	
6	2.85		0.000	7601	66435	2	
8	3.05		0.000	7072	60816	2	
9	3.21		0.000	5904	69131	2	
25	7.17	fluoride	10.534	1383385	21622570	3	0.00
27	8.24		0.000	10896	80628	2	
28	8.37	acetate	0.146	17549	292284	2	0.00
29	9.09	formate	0.431	15181	295995	2	0.00
30	9.84		0.000	5048	97886	2	
Totals			11.111	1476324	22874491		

File: 99012201.D06 Sample: S98T002932

Data Reprocessed On 01/25/1999 10:41:03

```

=====
Sample Name: S98T002932                      Date: 01/22/1999 13:34:14
Data File  : E:\DATA\99012201.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      21    5017    5Hz    6.73  10.11    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.70		0.000	3787	86535	2	
3	1.70		0.000	19901	202212	1	
6	2.85		0.000	7601	66447	2	
8	3.05		0.000	7072	61088	2	
9	3.21		0.000	5904	69131	2	
25	7.17	fluoride	10.535	1383385	21623909	3	0.00
27	8.37	acetate	0.334	17549	381325	2	0.00
28	9.09	formate	0.431	15181	295995	2	0.00
29	9.84		0.000	5048	97886	2	

```

-----
Totals                      11.299    1465427    22884527
=====

```

File: 99012201.D06 Sample: S98T002932

```

=====
Sample Name: S98T002932 DUP          Date: 01/22/1999 13:52:06
Data File  : C:\DX\DATA\99012201.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7   Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    4880  5Hz   0.00 16.26    50000

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.70		0.000	14817	139120	1	
4	2.75		0.000	5885	58000	2	
6	3.05		0.000	7256	86737	2	
7	3.33		0.000	6968	98670	2	
8	3.60		0.000	5232	66968	2	
10	4.00		0.000	5787	76717	2	
11	4.27		0.000	5731	107513	2	
22	7.17	fluoride	11.186	1529129	23002078	1	0.00
24	8.35	acetate	0.075	15870	258770	1	0.00
26	9.09	formate	0.304	13428	195128	1	0.00
Totals			11.566	1610101	24089701		

File: 99012201.D07 Sample: S98T002932 DUP

Data Reprocessed On 01/25/1999 10:42:02

Sample Name: S98T002932 DUP	Date: 01/22/1999 13:52:06
Data File : E:\DATA\99012201.D07	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 7	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	4880	5Hz	6.43	9.75	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.70		0.000	14817	139120	1	
4	2.75		0.000	5885	58000	2	
6	3.05		0.000	7256	86737	2	
7	3.33		0.000	6968	98670	2	
8	3.60		0.000	5232	66968	2	
10	4.00		0.000	5787	76717	2	
11	4.27		0.000	5731	107513	2	
22	7.17	fluoride	11.186	1529129	23002078	1	0.00
24	8.35	acetate	0.075	15870	258770	1	0.00
26	9.09	formate	0.304	13428	195128	1	0.00
Totals			11.566	1610101	24089701		

File: 99012201.D07 Sample: S98T002932 DUP

```

=====
Sample Name: S98T002932 SPIKE          Date: 01/22/1999 14:11:07
Data File  : C:\DX\DATA\99012201.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8      Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21  5029  5Hz  0.00 16.76      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	17700	173544	1	
5	2.87		0.000	9202	179459	2	
6	3.03		0.000	9399	102981	2	
7	3.19		0.000	9155	109527	2	
8	3.51		0.000	7768	131343	2	
9	3.97		0.000	8966	130049	2	
10	4.07		0.000	19441	79295	2	
11	4.24		0.000	8652	134821	2	
12	4.51		0.000	5658	51197	2	
13	4.71		0.000	4803	59835	2	
15	5.07		0.000	6422	67725	2	
16	5.29		0.000	5739	95068	2	
21	7.17	fluoride	25.809	3607345	54574885	2	0.00
22	7.87	glycolate	22.792	519755	9763432	2	0.00
23	8.35	acetate	20.015	494185	9676959	2	0.00
24	9.09	formate	24.718	1020665	19977997	2	0.00
Totals			93.334	5754855	95308120		

File: 99012201.D08 Sample: S98T002932 SPIKE

```

=====
Sample Name: S98T002932 SPIKE                      Date: 01/22/1999 14:11:07
Data File  : E:\DATA\99012201.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8                 Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5029  5Hz   6.11  10.90      50000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	17700	173544	1	
5	2.87		0.000	9202	179459	2	
6	3.03		0.000	9399	102981	2	
7	3.19		0.000	9155	109527	2	
8	3.51		0.000	7768	131343	2	
9	3.97		0.000	8966	130049	2	
10	4.07		0.000	19441	79295	2	
11	4.24		0.000	8652	134821	2	
12	4.51		0.000	5658	51197	2	
13	4.71		0.000	4803	59835	2	
15	5.07		0.000	6422	67725	2	
16	5.29		0.000	5739	95068	2	
21	7.17	fluoride	25.809	3607345	54574885	2	0.00
22	7.87	glycolate	22.792	519755	9763432	2	0.00
23	8.35	acetate	20.015	494185	9676959	2	0.00
24	9.09	formate	24.718	1020665	19977997	2	0.00
Totals			93.334	5754855	95308120		

LABCORE Completed Worklist Report for Worklist# 27736

Analyst: adp

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 FOR @IC4G-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0 @IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0 @IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0 @IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0 @IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0 @IC4G-QC F*4	QC	5.89e1	6.22e+01	105.603 % Recovery	
2	LCS-INST	0 @IC4G-QC ACETATE2	QC	1.24e2	1.14e+02	91.935 % Recovery	
2	LCS-INST	0 @IC4G-QC FORMATE2	QC	1.19e2	1.12e+02	94.119 % Recovery	
2	LCS-INST	0 @IC4G-QC GLYCOLTI	QC	1.07e2	1.03e+02	96.262 % Recovery	
3	CCV	0 @IC4G-QC F*4	QC	6.21e1	6.06e+01	97.585 % Recovery	
3	CCV	0 @IC4G-QC ACETATE2	QC	1.23e2	1.19e+02	96.748 % Recovery	
3	CCV	0 @IC4G-QC FORMATE2	QC	1.30e2	1.27e+02	97.692 % Recovery	
3	CCV	0 @IC4G-QC GLYCOLTI	QC	1.02e2	9.99e+01	97.941 % Recovery	
4	BLNK-PREP	0 @IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4	BLNK-PREP	0 @IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5	SAMPLE	S98T002933 0 W @IC4G-01 F*4-01	SOLID	N/A	8.130e+02	36.850 ug/g	
5	SAMPLE	S98T002933 0 W @IC4G-01 ACETATE2	SOLID	N/A	5.944e+01	23.780 ug/g	
5	SAMPLE	S98T002933 0 W @IC4G-01 FORMATE2	SOLID	N/A	7.371e+01	73.680 ug/g	
5	SAMPLE	S98T002933 0 W @IC4G-01 GLYCOLTI	SOLID	N/A	6.657e+01	66.600 ug/g	
6	DUP	S98T002933 0 W @IC4G-01 F*4-01	SOLID	8.13e+02	7.70e+02	5.433 RPD	
6	DUP	S98T002933 0 W @IC4G-01 ACETATE2	SOLID	5.94e+01	6.91e+01	15.097 RPD	
6	DUP	S98T002933 0 W @IC4G-01 FORMATE2	SOLID	<7.37e1	<7.39e1	RPD	
6	DUP	S98T002933 0 W @IC4G-01 GLYCOLTI	SOLID	<6.66e1	<6.68e1	RPD	
7	SPK	S98T002933 0 W @IC4G-01 F*4-01	SOLID	5.89e1	6.89e+01	116.978 % Recovery	
7	SPK	S98T002933 0 W @IC4G-01 ACETATE2	SOLID	1.24e2	1.15e+02	92.742 % Recovery	
7	SPK	S98T002933 0 W @IC4G-01 FORMATE2	SOLID	1.19e2	1.13e+02	94.958 % Recovery	
7	SPK	S98T002933 0 W @IC4G-01 GLYCOLTI	SOLID	1.07e2	9.97e+01	93.178 % Recovery	
8	SAMPLE	S98T002934 0 W @IC4G-01 F*4-01	SOLID	N/A	8.380e+02	36.340 ug/g	
8	SAMPLE	S98T002934 0 W @IC4G-01 FORMATE2	SOLID	N/A	2.057e+02	72.660 ug/g	
8	SAMPLE	S98T002934 0 W @IC4G-01 GLYCOLTI	SOLID	N/A	6.565e+01	65.640 ug/g	
9	DUP	S98T002934 0 W @IC4G-01 F*4-01	SOLID	8.38e+02	7.59e+02	9.894 RPD	
9	DUP	S98T002934 0 W @IC4G-01 ACETATE2	SOLID	3.79e+02	3.06e+02	21.314 RPD	
9	DUP	S98T002934 0 W @IC4G-01 FORMATE2	SOLID	2.06e+02	1.83e+02	11.825 RPD	
9	DUP	S98T002934 0 W @IC4G-01 GLYCOLTI	SOLID	<6.57e1	<6.61e1	RPD	
10	SAMPLE	S98T002935 0 W @IC4G-01 GLYCOLTI	SOLID	N/A	6.774e+01	67.740 ug/g	
11	DUP	S98T002935 0 W @IC4G-01 F*4-01	SOLID	8.08e+02	1.13e+03	33.230 RPD	
11	DUP	S98T002935 0 W @IC4G-01 ACETATE2	SOLID	3.66e+02	4.84e+02	27.765 RPD	
11	DUP	S98T002935 0 W @IC4G-01 FORMATE2	SOLID	1.81e+02	2.49e+02	31.628 RPD	
11	DUP	S98T002935 0 W @IC4G-01 GLYCOLTI	SOLID	<6.77e1	<6.76e1	RPD	
12	SPK	S98T002935 0 W @IC4G-01 F*4-01	SOLID	5.89e1	6.96e+01	118.166 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27736

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
12	SPK	S98T002935	0	W	@IC4G-01 ACETATE2	SOLID	1.24e2	1.16e+02	93.548 %	Recovery
12	SPK	S98T002935	0	W	@IC4G-01 FORMATE2	SOLID	1.19e2	1.12e+02	94.118 %	Recovery
12	SPK	S98T002935	0	W	@IC4G-01 GLYCOLYT	SOLID	1.07e2	1.01e+02	94.393 %	Recovery

Final page for worklist# 27736

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 1/18/99

12/30/98 14:54
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27736

Analyst: ADP Instrument: IC 4552 Book# 35N21-C LCS
 Method: LA-533-105 Rev/Mod F-O 34N21-A CCV
 Worklist Comment: BY-105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002933 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
6 DUP	S98T002933 0 W		@IC4G-01	SOLID		
7 SPK	S98T002933 0 W		@IC4G-01	SOLID		
8 SAMPLE	S98T002934 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
9 DUP	S98T002934 0 W		@IC4G-01	SOLID		
10 SAMPLE	S98T002935 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
11 DUP	S98T002935 0 W		@IC4G-01	SOLID		
12 SPK	S98T002935 0 W		@IC4G-01	SOLID		

Final page for worklist # 27736

Anthony Purieth 1-11-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 1-14-99
J. Howell
27736 JAN. CSV

S98T002934 & 2935 rejected
for same analytes due to
High RAD. Rest validated

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Reanalyzed on wk 27935
1/18/99
for 3/24/99

Data Reprocessed On 01/11/1999 00:00:10

Sample Name: 35N21-C LCS Date: 01/10/1999 23:57:59
 Data File : C:\DX\DATA\99011001.D15
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 15 Detector: CDM-2
 Analyst : *Anthony Paruto* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	6310	123352	1	
6	3.20		0.000	2205	32754	1	
9	4.01		0.000	1538	25705	2	
13	4.92		0.000	2530	23641	2	
14	5.05		0.000	2676	18135	2	
15	5.19		0.000	2522	16890	2	
16	5.32		0.000	5068	49122	2	
17	5.51		0.000	3809	38497	2	
18	5.73		0.000	2616	35380	2	
19	6.04		0.000	3574	40841	2	
20	6.25		0.000	2912	23265	2	
21	6.39		0.000	3918	71061	2	
22	7.31	fluoride	62.246	1848383	26729104	2	0.00
23	8.00	glycolate	102.988	510012	9169410	2	0.00
24	8.51	acetate	114.096	595059	11421571	2	0.00
25	9.20	formate	112.315	997088	18855315	2	0.00
26	11.31		0.000	10399577	453833651	2	
27	12.05		0.000	9548983	722331159	2	

Totals 391.645 23938780 1242838852

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 1525 TO 1541.
 File: 99011001.D15 Sample: 35N21-C LCS

Data Reprocessed On 01/11/1999 00:33:46

```

=====
Sample Name: 34N21-A CCV                               Date: 01/11/1999 00:17:40
Data File   : C:\DX\DATA\99011001.D16
Method      : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16                     Detector: CDM-2
Analyst     :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700 5Hz   0.00 19.00    16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
6	2.71		0.000	2475	22407	2	
8	3.12		0.000	1940	23835	2	
10	3.67		0.000	1717	16223	2	
15	4.87		0.000	1273	21230	2	
16	5.27		0.000	1881	16830	2	
18	5.55		0.000	2814	19417	2	
19	5.68		0.000	3397	43251	2	
20	6.07		0.000	2636	47713	2	
21	6.27		0.000	2800	25984	2	
22	6.49		0.000	3820	39281	2	
23	7.28	fluoride	60.554	1799717	25980856	2	0.00
24	7.97	glycolate	99.937	489726	8896643	2	0.00
25	8.48	acetate	119.497	619360	11949551	2	0.00
26	9.20	formate	126.727	1144265	21321923	2	0.00
27	11.31		0.000	10380694	448748940	2	
28	12.05		0.000	9364622	720396418	2	

```

-----
Totals          406.714    23823138 1237570502
-----

```

Sample Name: BLANK

Date: 01/10/1999 23:37:55

Data File : C:\DX\DATA\99011001.D14

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 14

Detector: CDM-2

Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	6279	210738	2	
6	2.89		0.000	3484	49565	2	
21	5.24		0.000	2232	29036	2	
28	6.72		0.000	2103	17692	1	
30	7.25	fluoride	0.008	1664	17256	2	0.00
36	9.23	formate	0.007	4565	75143	2	0.00
39	11.31		0.000	10414222	455367252	2	
40	12.03		0.000	9055985	708300058	2	
Totals			0.016	19490535	1164066740		

File: 99011001.D14 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 01/11/1999 00:51:17
Data File  : C:\DX\DATA\99011001.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17         Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	3080	80600	1	
11	9.20		0.000	494178	126393461	2	
12	9.71		0.000	467244	12902770	2	
13	11.31		0.000	11051102	574721392	2	
14	12.03		0.000	11065723	691772968	2	
Totals			0.000	23081326	1405871191		

File: 99011001.D17 Sample: PREP BLANK

Data Reprocessed On 01/14/1999 11:16:35

```

=====
Sample Name: S98T002933 SAM                      Date: 01/11/1999 01:32:11
Data File  : E:\DATA\99011001.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   6.31 10.96    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.27		0.000	3616	84354	1	
3	1.70		0.000	148945	1853862	1	
4	3.01		0.000	15820	513629	2	
5	3.28		0.000	13970	65933	2	
6	3.40		0.000	12967	124419	2	
7	3.55		0.000	12399	78539	2	
8	3.67		0.000	11736	92461	2	
9	3.84		0.000	10444	110027	2	
10	4.27		0.000	9583	275371	2	
11	4.61		0.000	7117	44213	2	
12	4.77		0.000	6747	92410	2	
13	4.96		0.000	5351	72479	2	
15	5.35		0.000	3258	30884	2	
21	6.59		0.000	2797	22054	2	
22	7.17	fluoride	4.103	1840407	29748427	2	0.00
23	8.35	acetate	0.300	32698	720793	2	0.00
24	9.07	formate	0.242	31746	627742	2	0.00
25	9.81		0.000	8300	166482	2	
26	11.15		0.000	15378115	320426445	2	
27	11.57		0.000	52632124710231596315		2	
Totals			4.645	54387726410586746839			

Data Reprocessed On 01/14/1999 11:15:04

```

=====
Sample Name: S98T002933 DUP                      Date: 01/11/1999 01:51:51
Data File  : E:\DATA\99011001.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   6.74 10.66    16000

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	5637	121488	1	
2	1.70		0.000	149810	1840900	1	
3	2.89		0.000	9782	180253	2	
4	3.01		0.000	10053	143238	2	
5	3.33		0.000	5875	82060	3	
10	4.36		0.000	3308	31340	2	
12	4.71		0.000	2638	21874	2	
13	4.91		0.000	3252	23746	2	
14	5.07		0.000	2623	20076	2	
15	5.27		0.000	3597	33347	2	
16	5.44		0.000	3362	36193	2	
17	5.65		0.000	4790	37208	2	
18	5.79		0.000	4552	56375	2	
22	7.17	fluoride	3.874	1715070	28040247	2	0.00
23	8.37	acetate	0.348	34554	800235	2	0.00
24	9.07	formate	0.263	34427	687951	2	0.00
25	9.84		0.000	9660	150356	1	
26	10.91		0.000	2617213	35971065	2	
27	11.15		0.000	15181774	271768460	2	
28	11.57		0.000	52232477110158486039		2	
Totals			4.485	54212674710498532452			

Data Reprocessed On 01/14/1999 11:13:19

```

=====
Sample Name: S98T002933 SPK                      Date: 01/11/1999 02:11:38
Data File  : E:\DATA\99011001.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 21              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6 5700 5Hz 6.05 10.93      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	7431	181536	1	
2	1.70		0.000	145836	1770646	1	
3	3.03		0.000	15218	347115	2	
4	3.11		0.000	15099	110260	2	
5	3.24		0.000	14507	102234	2	
6	3.39		0.000	14081	259821	2	
7	3.79		0.000	11285	144938	2	
8	3.93		0.000	10799	69190	2	
9	4.13		0.000	10741	91069	2	
10	4.25		0.000	10862	69929	2	
11	4.45		0.000	11024	121914	2	
12	4.53		0.000	10785	95242	2	
13	4.83		0.000	11714	184585	2	
14	5.15		0.000	11897	174116	2	
15	5.27		0.000	11438	72594	2	
16	5.39		0.000	10179	65622	2	
17	5.60		0.000	10007	112611	2	
18	5.71		0.000	10062	65118	2	
19	5.84		0.000	10409	43557	2	
20	5.91		0.000	10463	127086	2	
21	6.25		0.000	5860	47118	2	
22	7.17	fluoride	7.546	3408418	55904073	2	0.00
23	7.89	glycolate	4.983	394564	7467807	2	0.00
24	8.37	acetate	6.066	499928	10248438	2	0.00
25	9.07	formate	5.635	802321	15879110	3	0.00
26	9.84		0.000	7151	124348	4	
27	10.91		0.000	2618354	36205339	2	
28	11.15		0.000	14920965	270555105	2	
29	11.60		0.000	5417211	1710164164502	2	
Totals			24.230	56474251310564805024			

Data Reprocessed On 01/14/1999 11:18:52

```

=====
Sample Name: S98T002934 SAM                      Date: 01/11/1999 02:35:02
Data File  : E:\DATA\99011001.D22
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 22              Detector::CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   6.22 10.79   16000

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	3424	67384	1	
3	1.70		0.000	144787	1834812	2	
4	3.05		0.000	14511	366299	2	
5	3.21		0.000	14725	140209	2	
6	3.33		0.000	14526	205260	2	
7	3.63		0.000	13699	205036	2	
8	3.87		0.000	13761	67614	2	
9	3.93		0.000	13650	158409	2	
10	4.16		0.000	12634	138355	2	
11	4.43		0.000	12991	382226	2	
12	4.96		0.000	11026	128455	2	
13	5.17		0.000	10352	191919	2	
14	5.53		0.000	8635	127485	2	
15	5.76		0.000	8199	151270	2	
17	7.15	fluoride	4.289	1925470	31140971	2	0.00
18	8.35	acetate	1.940	151527	3437138	2	0.00
19	9.04	formate	1.053	149120	2898352	2	0.00
20	9.79		0.000	31326	626486	2	
21	10.85		0.000	5043035	53871756	2	
22	11.15		0.000	37765746	448165037	2	
23	11.60		0.000	480673481	9011025061	2	
24	18.56		0.000	1352	17680	1	
Totals			7.282	526037978	9555347215		

Data Reprocessed On 01/14/1999 11:20:46

Sample Name: S98T002934 DUP Date: 01/11/1999 02:54:39
 Data File : E:\DATA\99011001.D23
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 23 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 6 5700 5Hz 5.96 10.79 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	5804	129840	1	
2	1.70		0.000	135246	1614148	1	
3	3.03		0.000	12840	319849	2	
4	3.16		0.000	11760	96640	2	
5	3.28		0.000	10860	94816	2	
6	3.51		0.000	9080	176398	2	
7	3.83		0.000	5140	57722	2	
8	4.07		0.000	3740	30718	2	
9	4.24		0.000	2660	18474	2	
18	5.89		0.000	2271	31897	2	
19	7.17	fluoride	3.860	1726970	27938775	3	0.00
20	7.95	glycolate	0.026	15457	206213	4	0.00
21	8.37	acetate	1.554	132156	2797840	2	0.00
22	9.09	formate	0.932	132368	2559899	2	0.00
23	9.84		0.000	23848	438464	2	
24	10.88		0.000	4898373	52038766	2	
25	11.15		0.000	35101447	396949596	2	
26	11.60		0.000	481663572	9076597775	2	

Totals 6.372 523893592 9562097830

File: 99011001.D23 Sample: S98T002934 DUP

Data Reprocessed On 01/14/1999 11:21:55

```

=====
Sample Name: S98T002935 SAM                      Date: 01/11/1999 03:14:52
Data File  : E:\DATA\99011001.D24
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 24
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   5.81 10.84   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	5806	159096	1	
2	1.70		0.000	110950	1321168	1	
3	3.08		0.000	6634	278920	2	
4	3.29		0.000	2706	18046	3	
15	5.36		0.000	2818	50694	2	
16	5.68		0.000	4076	21860	2	
17	5.77		0.000	4496	20895	2	
18	5.84		0.000	5165	81665	2	
20	7.20	fluoride	4.010	1877227	29053804	2	0.00
21	8.37	acetate	1.813	146897	3226855	2	0.00
22	9.09	formate	0.898	128274	2463278	2	0.00
23	9.87		0.000	22898	432593	2	
24	10.88		0.000	4949803	52772244	2	
25	11.15		0.000	34677610	426664686	2	
26	11.65		0.000	419112144	6960149493	3	
27	12.27		0.000	6728392	404616110	4	
Totals			6.720	467785893	7881331407		

File: 99011001.D24 Sample: S98T002935 SAM

Data Reprocessed On 01/14/1999 11:23:25

```

=====
Sample Name: S98T002935 DUP                      Date: 01/11/1999 03:38:34
Data File  : E:\DATA\99011001.D25
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 25              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          6    5700  5Hz   6.03 10.88    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	0.93		0.000	1197	18038	1	
4	1.70		0.000	111772	1409074	1	
5	3.00		0.000	9626	328948	2	
6	3.43		0.000	4611	63557	2	
7	3.65		0.000	1636	21902	2	
11	4.40		0.000	1797	37611	2	
13	4.92		0.000	3513	50966	2	
14	5.31		0.000	4682	53138	2	
15	5.53		0.000	5213	99169	2	
16	5.80		0.000	6469	114952	2	
17	7.17	fluoride	5.606	2712243	41058630	2	0.00
18	8.35	acetate	2.408	197722	4210415	2	0.00
19	9.04	formate	1.241	180655	3426014	2	0.00
20	9.79		0.000	34589	679269	2	
21	10.85		0.000	6175077	60231978	2	
22	11.15		0.000	42048339	525645818	2	
23	11.71		0.000	346097746	5476834479	3	
24	12.24		0.000	9613556	414454740	4	

```

-----
Totals          9.254  407210442  6528738698
-----

```

File: 99011001.D25 Sample: S98T002935 DUP

1539

Minutes

Data Reprocessed On 01/14/1999 11:25:02

```

=====
Sample Name: S98T002935 SPK                      Date: 01/11/1999 03:58:51
Data File  : E:\DATA\99011001.D26
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 26              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   6.35 10.84   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	119960	1411926	2	
4	2.99		0.000	11246	309352	2	
5	3.12		0.000	9902	78516	2	
6	3.25		0.000	8858	73578	2	
7	3.41		0.000	6406	36391	2	
8	3.55		0.000	5522	42116	2	
9	3.71		0.000	3930	42766	2	
10	3.88		0.000	2413	16882	2	
19	5.31		0.000	1907	16790	2	
20	5.56		0.000	1673	18802	2	
23	7.15	fluoride	7.489	3458335	55465167	2	0.00
24	7.87	glycolate	5.050	409676	7567344	2	0.00
25	8.35	acetate	7.609	634236	12788171	2	0.00
26	9.07	formate	6.504	966076	18370764	3	0.00
27	9.81		0.000	23199	426524	4	
28	10.88		0.000	5003070	52701715	2	
29	11.15		0.000	35134974	430735240	2	
30	11.65		0.000	420309693	6952453609	3	
31	12.27		0.000	6859785	402658693	4	
32	18.72		0.000	1600	17600	4	

```

-----
Totals                26.651  472972463  7935231947

```

LABCORE Completed Worklist Report for Worklist# 27751

Analyst: adp

Instrument: IC45S2

Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 @IC4G-01 RERUN RTS (USE DF 5151)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0	@IC4G-QC F*4	QC	5.69e1	6.20e+01	105.263 % Recovery	
2	LCS-INST	0	@IC4G-QC ACETATE2	QC	1.24e2	1.14e+02	91.935 % Recovery	
2	LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.13e+02	94.958 % Recovery	
2	LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.07e2	1.03e+02	96.262 % Recovery	
3	CCV	0	@IC4G-QC F*4	QC	6.21e1	6.00e+01	96.618 % Recovery	
3	CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.18e+02	95.935 % Recovery	
3	CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.25e+02	96.154 % Recovery	
3	CCV	0	@IC4G-QC GLYCOLT1	QC	1.02e2	9.72e+01	95.294 % Recovery	
4	SAMPLE	S98T003384	@IC4G-01 F*4-01	LIQUID	N/A	2.247e+02	159.700	ug/mL
4	SAMPLE	S98T003384	@IC4G-01 ACETATE2	LIQUID	N/A	1.406e+03	298.800	ug/mL
4	SAMPLE	S98T003384	@IC4G-01 FORMATE2	LIQUID	N/A	9.795e+02	319.400	ug/mL
5	DUP	S98T003384	@IC4G-01 F*4-01	LIQUID	2.25e+02	2.26e+02	0.443	RPD
5	DUP	S98T003384	@IC4G-01 ACETATE2	LIQUID	1.41e+03	1.54e+03	8.814	RPD
5	DUP	S98T003384	@IC4G-01 FORMATE2	LIQUID	9.80e+02	1.02e+03	4.000	RPD
5	DUP	S98T003384	@IC4G-01 GLYCOLT1	LIQUID	?	<2.88e2		RPD
6	SPK	S98T003384	@IC4G-01 F*4-01	LIQUID	5.89e1	6.12e+01	103.905 % Recovery	
6	SPK	S98T003384	@IC4G-01 ACETATE2	LIQUID	1.24e2	1.21e+02	97.581 % Recovery	
6	SPK	S98T003384	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.17e+02	98.319 % Recovery	
6	SPK	S98T003384	@IC4G-01 GLYCOLT1	LIQUID	1.07e2	0.00e+00	0.000 % Recovery	

Final page for worklist# 27751

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

James M. Lacy 1/18/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27751

Analyst: ADP Instrument: IC 4552 Book# 3SN21-C LCS
Method: LA-533-105 Rev/Mod F-0 34N21-A CCV
Worklist Comment: BY105 @IC4G-01 RERUN RTS (USE DF 5151)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T003384 0		@IC4G-01 LIQUID		98000437	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2					
5 DUP	S98T003384 0		@IC4G-01 LIQUID			
6 SPK	S98T003384 0		@IC4G-01 LIQUID			

Final page for worklist # 27751

Anthony Purnell
Analyst Signature Date 1-10-99

_____
Analyst Signature Date

Data Entry Comments:

uploaded 1-14-99
John W. Wall
27751 TAN.CSV
reuploaded to correct spk value left out
1/18/99 Jm. Lugo
Validated 1/18/99 Jm. Lugo

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Data Reprocessed On 01/10/1999 22:00:08

```

=====
Sample Name: 35N21-C LCS                               Date: 01/10/1999 21:57:23
Data File  : C:\DX\DATA\99011001.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9                     Detector: CDM-2
Analyst    : Antony Parra Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.63		0.000	2548	100511	2	
4	2.28		0.000	2302	23752	1	
5	2.73		0.000	3909	37846	2	
6	2.81		0.000	4217	26734	2	
7	3.00		0.000	4771	55240	2	
9	3.35		0.000	4203	32359	2	
10	3.68		0.000	2577	25066	2	
20	5.41		0.000	4753	43555	2	
21	5.69		0.000	2270	29280	2	
22	5.95		0.000	1222	19407	2	
26	7.28	fluoride	61.978	1857208	26610599	2	0.00
27	7.97	glycolate	103.023	510826	9172540	2	0.00
28	8.48	acetate	114.115	593705	11423407	2	0.00
29	9.17	formate	112.788	999522	18935987	2	0.00
30	11.31		0.000	10458662	459223622	2	
31	12.05		0.000	9686161	726781852	2	
Totals			391.903	24138858	1252541757		

File: 99011001.D09 Sample: 35N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1544 TO 1557.

Data Reprocessed On 01/10/1999 22:21:38

```

=====
Sample Name: 34N21-A A CCV                      Date: 01/10/1999 22:17:39
Data File  : C:\DX\DATA\99011001.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          101    5700    5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	4911	195052	1	
5	2.79		0.000	2649	19259	2	
6	2.88		0.000	4439	34259	2	
7	3.12		0.000	2984	36578	2	
21	5.84		0.000	1676	27692	2	
22	6.12		0.000	1425	18919	2	
25	7.31	fluoride	59.965	1788581	25720742	2	0.00
26	8.00	glycolate	97.235	480349	8655476	2	0.00
27	8.51	acetate	118.022	614337	11805412	2	0.00
28	9.23	formate	124.588	1133230	20955101	2	0.00
29	11.31		0.000	10420344	456925521	2	
30	12.03		0.000	9224765	715187110	2	
Totals			399.810	23679689	1239581121		

```

=====
Sample Name: BLANK                      Date: 01/10/1999 21:36:53
Data File  : C:\DX\DATA\99011001.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8
Analyst    :                          Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	2473	21692	1	
4	1.50		0.000	3837	68339	2	
17	3.93		0.000	1964	18660	2	
25	5.49		0.000	2581	19806	2	
26	5.65		0.000	2708	18809	2	
27	5.85		0.000	1983	18431	2	
30	6.43		0.000	2382	16947	2	
34	7.41	fluoride	0.008	1315	16847	2	0.00
36	8.00		0.000	2007	36976	2	
37	8.43		0.000	1348	17386	2	
42	11.31		0.000	10999984	494419425	2	
43	12.03		0.000	9666022	744181693	2	
Totals			0.008	20688604	1238855010		

File: 99011001.D08 Sample: BLANK

Data Reprocessed On 01/12/1999 15:39:51

Sample Name: S98T003384 SAM	Date: 01/10/1999 22:38:15
Data File : E:\DATA\99011001.D11	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 11	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	5151	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	6318	234362	2	
3	2.25		0.000	5937	101870	1	
4	2.87		0.000	5827	98493	2	
5	3.07		0.000	5971	68797	2	
6	3.20		0.000	4761	69879	2	
19	5.41		0.000	2242	27672	2	
20	5.63		0.000	3711	105617	2	
21	6.12		0.000	3335	33395	2	
22	7.28	fluoride	224.721	93274	1539394	2	0.00
23	8.00	glycolate	-8.435	12181	154998	2	0.00
24	8.48	acetate	1405.862	148568	2936368	2	0.00
25	9.20	formate	979.450	170025	3145538	2	0.00
26	9.97		0.000	25872	412084	1	
27	10.88		0.000	5009641	55707109	2	
28	11.17		0.000	37344263	551160274	2	
29	11.89		0.000	70524787	1402397050	2	

Totals 2601.599 113366713 2018192900

File: 99011001.D11 Sample: S98T003384 SAM

Data Reprocessed On 01/12/1999 15:40:47

```

=====
Sample Name: S98T003384 DUP                      Date: 01/10/1999 22:57:55
Data File  : E:\DATA\99011001.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700   5Hz   0.00  19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.37		0.000	571	19184	1	
4	2.21		0.000	4141	62208	1	
5	2.84		0.000	5469	73138	2	
6	3.05		0.000	5798	120668	2	
7	3.43		0.000	3439	56499	2	
8	3.84		0.000	2024	18181	2	
13	4.69		0.000	1721	16779	2	
17	5.44		0.000	2807	33655	2	
18	5.60		0.000	3433	46071	2	
19	5.89		0.000	5529	31691	2	
20	5.96		0.000	6013	89487	2	
21	6.75		0.000	3792	48387	2	
22	7.25	fluoride	225.975	95810	1549940	2	0.00
23	8.45	acetate	1541.052	150893	3197026	2	0.00
24	9.17	formate	1021.164	175902	3281883	2	0.00
25	9.95		0.000	26832	440422	1	
26	10.88		0.000	5140965	61914967	2	
27	11.17		0.000	38828176	546371325	2	
28	11.89		0.000	73110148	1407568514	2	
Totals			2788.191	117573463	2024940025		

```

=====
Sample Name: S98T003384 SPK                      Date: 01/10/1999 23:17:47
Data File  : E:\DATA\99011001.D13
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 13              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          5151   5700   5Hz   0.00 19.00      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	2105	21032	1	
3	0.93		0.000	945	16904	1	
6	2.25		0.000	4060	66352	1	
8	2.69		0.000	2380	19297	2	
9	2.91		0.000	2346	28293	2	
13	3.53		0.000	1793	21360	2	
22	5.05		0.000	3273	37350	2	
23	5.21		0.000	3935	25989	2	
24	5.36		0.000	3947	21375	2	
25	5.45		0.000	4658	50085	2	
26	5.68		0.000	6670	70843	2	
27	5.84		0.000	6612	144453	2	
28	7.25	fluoride	3314.731	1916905	27945519	2	0.00
29	7.97	glycolate	5271.111	520186	9202290	2	0.00
30	8.45	acetate	7504.331	749973	14649537	2	0.00
31	9.17	formate	6865.582	1238854	22676680	2	0.00
32	9.92		0.000	29241	502087	2	
33	10.88		0.000	5445111	57821178	2	
34	11.17		0.000	39205839	552032643	2	
35	11.89		0.000	74326626	1415661830	3	
36	18.45		0.000	953	18571	4	
Totals			22955.754	123476411	2101033670		

LABCORE Completed Worklist Report for Worklist# 27761

Analyst: adp

Instrument: IC40S1

Book#: 50N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1 CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	@IC-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	@IC-QC F	QC	5.92e1	6.24e+1	105.405 % Recovery	
2 LCS-INST	0	@IC-QC CL	QC	7.96e1	7.95e+1	99.874 % Recovery	
2 LCS-INST	0	@IC-QC NO2	QC	5.15e2	5.15e+2	100.000 % Recovery	
2 LCS-INST	0	@IC-QC BR	QC	5.86e2	5.71e+2	97.440 % Recovery	
2 LCS-INST	0	@IC-QC NO3	QC	5.92e2	5.61e+2	94.764 % Recovery	
2 LCS-INST	0	@IC-QC P04	QC	5.48e2	5.49e+2	100.182 % Recovery	
2 LCS-INST	0	@IC-QC S04	QC	6.38e2	6.31e+2	98.903 % Recovery	
2 LCS-INST	0	@IC-QC OXALATE2	QC	5.22e2	5.57e+2	106.705 % Recovery	
3 CCV	0	@IC-QC F	QC	6.40e1	6.40e+1	100.000 % Recovery	
3 CCV	0	@IC-QC CL	QC	8.95e1	8.78e+1	98.101 % Recovery	
3 CCV	0	@IC-QC NO2	QC	5.45e2	5.41e+2	99.266 % Recovery	
3 CCV	0	@IC-QC BR	QC	6.30e2	6.23e+2	98.889 % Recovery	
3 CCV	0	@IC-QC NO3	QC	7.00e2	6.88e+2	98.286 % Recovery	
3 CCV	0	@IC-QC P04	QC	6.35e2	6.41e+2	100.945 % Recovery	
3 CCV	0	@IC-QC S04	QC	7.00e2	6.81e+2	97.286 % Recovery	
3 CCV	0	@IC-QC OXALATE2	QC	5.27e2	5.60e+2	106.262 % Recovery	
4 BLNK-PREP	0	@IC-01 F-02	SOLID	1	<1.20e-2		ug/g
4 BLNK-PREP	0	@IC-01 CL-02	SOLID	1	<1.70e-2		ug/g
4 BLNK-PREP	0	@IC-01 NO2-02	SOLID	1	2.48e-01	0.248	ug/g
4 BLNK-PREP	0	@IC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4 BLNK-PREP	0	@IC-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4 BLNK-PREP	0	@IC-01 P04-02	SOLID	1	<1.20e-1		ug/g
4 BLNK-PREP	0	@IC-01 S04-02	SOLID	1	<1.38e-1		ug/g
4 BLNK-PREP	0	@IC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
6 DUP	S98T002613	0 W @IC-01 F-02	SOLID	5.39e+02	5.08e+02	5.922 RPD	
6 DUP	S98T002613	0 W @IC-01 CL-02	SOLID	<1.41e2	<1.38e2	RPD	
6 DUP	S98T002613	0 W @IC-01 NO2-02	SOLID	3.31e+03	3.11e+03	6.231 RPD	
6 DUP	S98T002613	0 W @IC-01 BR-02	SOLID	<1.04e3	<1.02e3	RPD	
6 DUP	S98T002613	0 W @IC-01 NO3-02	SOLID	6.28e+05	5.60e+05	11.448 RPD	
6 DUP	S98T002613	0 W @IC-01 P04-02	SOLID	2.72e+03	2.64e+03	2.985 RPD	
6 DUP	S98T002613	0 W @IC-01 S04-02	SOLID	1.24e+03	<1.12e3	RPD	
6 DUP	S98T002613	0 W @IC-01 OXALATE2	SOLID	1.87e+03	1.73e+03	7.778 RPD	
7 SAMPLE	S98T002615	0 W @IC-01 F-02	SOLID	N/A	3.181e+02	98.480	ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27761

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 SAMPLE	S98T002615	0 W	@IC-01	CL-02	SOLID	N/A	1.601e+03	139.500 ug/g
7 SAMPLE	S98T002615	0 W	@IC-01	NO2-02	SOLID	N/A	3.068e+04	886.400 ug/g
7 SAMPLE	S98T002615	0 W	@IC-01	BR-02	SOLID	N/A	1.026e+03	1026.000 ug/g
7 SAMPLE	S98T002615	0 W	@IC-01	NO3-02	SOLID	N/A	5.877e+04	1141.000 ug/g
7 SAMPLE	S98T002615	0 W	@IC-01	PO4-02	SOLID	N/A	2.483e+04	984.800 ug/g
8 DUP	S98T002615	0 W	@IC-01	F-02	SOLID	3.18e+02	<9.56e1	RPD
8 DUP	S98T002615	0 W	@IC-01	CL-02	SOLID	1.60e+03	1.71e+03	6.647 RPD
8 DUP	S98T002615	0 W	@IC-01	NO2-02	SOLID	3.07e+04	3.29e+04	6.918 RPD
8 DUP	S98T002615	0 W	@IC-01	BR-02	SOLID	<1.03e3	<9.95e2	RPD
8 DUP	S98T002615	0 W	@IC-01	NO3-02	SOLID	5.88e+04	5.33e+04	9.813 RPD
8 DUP	S98T002615	0 W	@IC-01	PO4-02	SOLID	2.49e+04	1.68e+04	38.462 RPD
8 DUP	S98T002615	0 W	@IC-01	SO4-02	SOLID	1.95e+03	2.54e+03	26.281 RPD
8 DUP	S98T002615	0 W	@IC-01	OXALATE2	SOLID	1.46e+04	1.31e+04	10.830 RPD

Final page for worklist# 27761

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 1/25/99

LABCORE Data Entry Template for Worklist# 27761

Analyst: ADP Instrument: IC 4051 Book# SIN21-B LCS
Method: LA-533-105 Rev/Mod F-0 SON 21-B CCV
Worklist Comment: BY105 FOR @IC-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T002613	0	W	@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
6	DUP	S98T002613	0	W	@IC-01	SOLID		
7	SAMPLE	S98T002615	0	W	@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02						
8	DUP	S98T002615	0	W	@IC-01	SOLID		

Final page for worklist # 27761

Analyst Signature Date 01-21-99

_____
Analyst Signature Date

Data Entry Comments:
uploaded 1-25-99
ADP
27761.JAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

Reanalyzed S98T002615
on W2 # 28021. 3/24/99

Rejected all of S98T002613 because
R/D was unusually high, so may
have dilution or contamination problem.
S98T002615 PO4 SO4 rejected
due to high KPD probably due to
non homogeneous sample,
AM Luyze
1/25/99

Sample Name: 51N21-B LCS Date: 01/20/1999 23:26:38
 Data File : C:\DX\DATA\99012011.D12
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 12 Detector: CDM-1
 Analyst : *Anthony P. Smith* Column: AG4A/AS4A anion column
 1-20-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	74	257	1	
2	0.99	fluoride	62.366/15.35	1902	9479	1	0.34
3	1.44	chloride	79.493/9.87	2365	9942	1	0.70
4	1.69	nitrite	515.377/100.07	8879	43600	1	-0.97
5	2.47	bromide	571.470/97.52	6062	33048	1	0.54
6	2.79	nitrate	561.074/94.78	6802	43790	1	0.60
7	4.03	phosphate	548.956/104.17	1486	18124	1	1.85
8	5.12	sulfate	631.167/98.93	5084	61245	1	2.20
9	6.61	oxalate	557.426/106.79	2237	38183	1	2.48

Totals 3527.329 34890 257669

File: 99012011.D12 Sample: 51N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1555 TO 1560.

```

=====
Sample Name: 50N21-B CCV                      Date: 01/20/1999 23:49:29
Data File  : C:\DK\DATA\99012011.D14
Method     : C:\DK\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 14           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	86	318	2	
2	0.99	fluoride	64.027	1906	9735	2	-0.34
3	1.44	chloride	87.790	2577	10950	1	0.70
4	1.69	nitrite	540.837	9200	45767	1	-0.97
5	2.47	bromide	623.470	6456	36095	1	0.54
6	2.78	nitrate	687.855	8100	53884	1	0.36
7	4.03	phosphate	640.861	1731	21237	1	1.85
8	5.12	sulfate	681.075	5664	66134	1	2.20
9	6.61	oxalate	559.749	2299	38341	1	2.48
Totals			3885.665	38019	282460		

File: 99012011.D14 Sample: 50N21-B CCV

```

=====
Sample Name: BLANK                      Date: 01/20/1999 23:37:49
Data File  : C:\DX\DATA\99012011.D13
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 13      Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.43	chloride	-0.014	44	290	1	0.23
2	1.70	nitrite	0.022	13	52	1	-0.58
3	2.83	nitrate	0.048	38	200	1	2.05
Totals			0.055	95	542		

```

=====
Sample Name: PREP BLANK          Date: 01/21/1999 00:01:09
Data File  : C:\DX\DATA\99012011.D15
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 15      Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    3000  5Hz   0.00 10.00         30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.27		0.000	55	540	1	
2	1.71	nitrite	0.248	349	1983	1	-0.19
3	2.25		0.000	13	78	1	
4	2.83	nitrate	0.056	51	265	1	2.29
Totals			0.303	469	2865		

File: 99012011.D15 Sample: PREP BLANK

```

=====
Sample Name: S98T002613 SAM                      Date: 01/21/1999 00:27:28
Data File  : C:\DX\DATA\99012011.D17
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 17              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	188	852	2	
2	1.00	fluoride	2.666	207	1007	2	1.01
3	1.45	chloride	-0.255	85	382	1	1.63
4	1.71	nitrite	16.363	628	3278	1	-0.19
5	2.67	nitrate	3107.381	81040	759558	1	-3.49
6	4.08	phosphate	13.444	96	1036	1	3.03
7	5.17	sulfate	6.113	150	1865	1	3.26
8	6.72	oxalate	9.252	144	2188	1	4.13
9	7.52		0.000	79	2447	1	
Totals			3154.965	82617	772613		

File: 99012011.D17 Sample: S98T002613 SAM

Data Reprocessed On 01/25/1999 09:05:48

```

=====
Sample Name: S98T002613  DUP                      Date: 01/21/1999 00:39:57
Data File  : F:\DATA\99012011.D18
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 18              Detector:CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          41    3000   5Hz   0.00 10.00          30
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	184	834	2	
2	0.98	fluoride	2.578	199	974	2	-1.01
3	1.43	chloride	-0.237	95	387	1	-0.23
4	1.69	nitrite	15.683	632	3136	1	-1.36
5	2.65	nitrate	2820.495	76871	666244	1	-0.13
6	4.05	phosphate	13.300	96	1025	1	2.19
7	5.12	sulfate	5.074	123	1622	1	2.20
8	6.67	oxalate	8.720	138	2101	1	3.31
Totals			2865.614	78337	676323		

```

=====
Sample Name: S98T002615 SAM                      Date: 01/21/1999 00:51:06
Data File  : C:\DX\DATA\99012011.D19
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 19              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 41 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	100	525	2	
2	1.02	fluoride	1.589	70	609	2	3.03
3	1.43	chloride	7.998	633	2780	1	-0.23
4	1.69	nitrite	153.242	6351	31877	1	-1.36
5	2.77	nitrate	293.556	8586	56705	1	0.12
6	4.03	phosphate	124.014	834	9981	1	1.68
7	5.12	sulfate	9.750	201	2716	1	2.20
8	6.67	oxalate	72.661	845	12641	1	3.31
Totals			662.812	17620	117835		

File: 99012011.D19 Sample: S98T002615 SAM

```

=====
Sample Name: S98T002615 DUP                      Date: 01/21/1999 01:02:10
Data File  : C:\DX\DATA\99012011.D20
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 20              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41  3000 5Hz  0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	96	501	2	
2	1.08		0.000	66	579	2	
3	1.43	chloride	8.784	667	3010	1	0.23
4	1.69	nitrite	169.557	6814	35292	1	-0.97
5	2.79	nitrate	274.218	8133	52898	1	0.60
6	4.04	phosphate	86.726	569	6946	1	2.02
7	5.12	sulfate	13.053	243	3489	1	2.20
8	6.67	oxalate	67.405	773	11773	1	3.31
Totals			619.743	17361	114488		

File: 99012011.D20 Sample: S98T002615 DUP

LABCORE Completed Worklist Report for Worklist# 27763

Analyst: ehc

Instrument: IC45S2

Book#: 58N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC4G-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	3.07e-01	0.307	ug/mL
1 CCB	0		@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0		@IC4G-QC F*4	QC	6.10e1	6.30e+01	103.279 %	Recovery
2 LCS-INST	0		@IC4G-QC ACETATE2	QC	1.16e2	1.16e+02	100.000 %	Recovery
2 LCS-INST	0		@IC4G-QC FORMATE2	QC	1.19e2	1.27e+02	106.723 %	Recovery
2 LCS-INST	0		@IC4G-QC GLYCOLT1	QC	1.08e2	1.09e+02	100.926 %	Recovery
3 CCV	0		@IC4G-QC F*4	QC	6.29e1	5.55e+01	88.235 %	Recovery
3 CCV	0		@IC4G-QC ACETATE2	QC	1.20e2	1.14e+02	95.000 %	Recovery
3 CCV	0		@IC4G-QC FORMATE2	QC	1.20e2	1.20e+02	100.000 %	Recovery
3 CCV	0		@IC4G-QC GLYCOLT1	QC	1.03e2	9.52e+01	92.427 %	Recovery
4 BLNK-PREP	0		@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002613	0 W	@IC4G-01 ACETATE2	SOLID	N/A	2.392e+02	84.920	ug/g
5 SAMPLE	S98T002613	0 W	@IC4G-01 FORMATE2	SOLID	N/A	< 2.633e+02	263.300	ug/g
5 SAMPLE	S98T002613	0 W	@IC4G-01 GLYCOLT1	SOLID	N/A	< 2.378e+02	237.700	ug/g
6 DUP	S98T002613	0 W	@IC4G-01 F*4-01	SOLID	5.88e+02	5.70e+02	3.109	RPD
6 DUP	S98T002613	0 W	@IC4G-01 ACETATE2	SOLID	2.39e+02	2.13e+02	11.504	RPD
6 DUP	S98T002613	0 W	@IC4G-01 FORMATE2	SOLID	<2.63e2	<2.58e2		RPD
6 DUP	S98T002613	0 W	@IC4G-01 GLYCOLT1	SOLID	<2.38e2	<2.33e2		RPD

Final page for worklist# 27763

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jane Lyle 3/12/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27763

Analyst: EHK Instrument: IC 4552 Book# 59N21-A

Method: LA-533-105 Rev/Mod F-0 58N21-A

Worklist Comment: BY105 FOR @IC4G-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002613	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
6 DUP	S98T002613	0 W	@IC4G-01	SOLID		

Final page for worklist # 27763

Signature _____ Date _____

Signature _____ Date _____

Data Entry Comments:

uploaded 3-8-99

John Wavell

27763FEB.CSV

F rejected, rest validated

JMZ 3/4/99

Reanalyzed on lot# 28542. of 3/24/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1564

01/04/99 14:20
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27763

Analyst:

PAL

Instrument: IC

Book# *58N21-A*

Method: LA-533-105 Rev/Mod

58N21-A

Worklist Comment: BY105 FOR @IC4G-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	LCS-INST				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002613	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
6	DUP	S98T002613	0	W	@IC4G-01	SOLID		
7	SAMPLE	S98T002614	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
8	DUP	S98T002614	0	W	@IC4G-01	SOLID		
9	SPK	S98T002614	0	W	@IC4G-01	SOLID		
10	SAMPLE	S98T002615	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
11	DUP	S98T002615	0	W	@IC4G-01	SOLID		

*ENC-2-A-99**NEED WORK LIST FOR THESE*

Final page for worklist # 27763

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 59N21-A LCS Date: 02/19/1999 08:57:19
 Data File : C:\DX\DATA\99021911.D22
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 22 Detector: CDM-2
 Analyst : *EM* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4485	5Hz	0.00	14.95	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	11799	241912	1	
2	2.27		0.000	2256	30203	2	
4	2.83		0.000	2412	26106	2	
5	3.07		0.000	1664	19038	2	
17	5.51		0.000	1443	27761	2	
22	7.33	fluoride	63.010	1908648	28772726	2	0.00
23	8.05	glycolate	109.266	529526	9936585	2	0.00
24	8.53	acetate	115.937	536045	10603111	2	0.00
25	9.28	formate	127.151	1008341	19574102	2	0.00
Totals			415.364	4002135	69231543		

File: 99021911.D22 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1566 TO 1571.

Sample Name: 58N21-A CCV

Date: 02/19/1999 09:15:05

Data File : C:\DX\DATA\99021911.D23

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 23

Detector: CDM-2

Analyst : *EW. Loh* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4483	5Hz	0.00	14.94	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	2218	25312	1	
13	3.44		0.000	350	18775	2	
30	7.31	fluoride	55.548	1631331	25198875	2	0.00
31	8.03	glycolate	95.215	457742	8608686	2	0.00
32	8.53	acetate	114.190	529731	10440934	2	0.00
33	9.25	formate	120.439	935243	18489071	2	0.00
Totals			385.392	3556615	62781653		

File: 99021911.D23 Sample: 58N21-A CCV

```

=====
Sample Name: INSTR BLANK                      Date: 02/19/1999 08:41:28
Data File  : C:\DX\DATA\99021911.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 21          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 4745 5Hz 0.00 15.81 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	53122	1409228	1	
2	1.30		0.000	26419	462640	1	
5	2.56		0.000	8585	79953	2	
6	2.67		0.000	9558	205012	2	
7	3.04		0.000	6372	126254	3	
11	4.15		0.000	2979	23793	2	
13	6.40		0.000	360867	29243439	1	
14	9.25	formate	0.307	12672	4215058	3	0.00
15	9.41		0.000	2407	40673	4	
Totals			0.307	482981	35806050		

File: 99021911.D21 Sample: INSTR BLANK

Sample Name: PREP BLANK Date: 02/19/1999 09:33:00
 Data File : C:\DX\DATA\99021911.D24
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 24 Detector: CDM-2
 Analyst : *PHC* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	4475	5Hz	0.00	14.91	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	1369	22172	1	
14	3.96		0.000	1262	17058	2	
27	7.39	fluoride	0.024	2357	35267	2	0.00
34	9.28	formate	0.047	3968	74660	2	0.00
Totals			0.070	8956	149158		

File: 99021911.D24 Sample: PREP BLANK

```

=====
Sample Name: S98T002613      Date: 02/19/1999 09:48:20
Data File  : C:\DX\DATA\99021911.D25
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 25      Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21  4309  5Hz  0.00 14.36   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.73		0.000	14240	118752	1	
6	2.92		0.000	2131	29643	2	
12	4.47		0.000	937	16792	1	
21	7.25	fluoride	2.908	342814	5478539	2	0.00
22	8.45	acetate	1.183	13322	295428	2	0.00
23	9.23	formate	1.172	12329	221058	2	0.00
Totals			5.263	385774	6160213		

File: 99021911.D25 Sample: S98T002613

HNF-1663 REV. 0

Sample Name: S98T002613 DUP Date: 02/19/1999 09:59:58
 Data File : C:\DX\DATA\99021911.D26
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 26 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

alibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
external	1	21	3197	5Hz	0.00	10.65	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.70		0.000	9708	71672	1	
2	2.68		0.000	4192	39284	2	
3	2.76		0.000	4526	130534	2	
15	5.87		0.000	2004	35062	2	
17	7.20	fluoride	2.875	335035	5403937	2	0.00
18	8.40	acetate	1.074	10372	246039	2	0.00
19	9.15	formate	1.176	10445	223958	2	0.00
20	9.97		0.000	1937	35280	1	
Totals			5.125	378219	6185766		

File: 99021911.D26 Sample: S98T002613 DUP

LABCORE Completed Worklist Report for Worklist# 27850

Analyst: adp Instrument: IC45S2 Book#: 35N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 @IC4G-01

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4 QC	QC	1	8.10e-02	0.081	ug/mL
1 CCB	0	@IC4G-QC ACETATE2 QC	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2 QC	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI QC	QC	1	5.40e-02	0.054	ug/mL
2 LCS-INST	0	@IC4G-QC F*4 QC	QC	5.89e1	6.25e+01	106.112 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2 QC	QC	1.24e2	1.14e+02	91.935 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2 QC	QC	1.19e2	1.14e+02	95.798 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLTI QC	QC	1.07e2	1.07e+02	100.000 % Recovery	
3 CCV	0	@IC4G-QC F*4 QC	QC	6.21e1	5.61e+01	90.338 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2 QC	QC	1.23e2	1.13e+02	91.870 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2 QC	QC	1.30e2	1.17e+02	90.000 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLTI QC	QC	1.02e2	9.25e+01	90.686 % Recovery	
4 BLNK-PREP	0	@IC4G-01 F*4-01 SOLID	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2 SOLID	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2 SOLID	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLTI SOLID	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T003400 0 W	@IC4G-01 F*4-01 SOLID	SOLID	N/A	1.056e+02	67.290	ug/g
5 SAMPLE	S98T003400 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	N/A	1.486e+03	43.410	ug/g
5 SAMPLE	S98T003400 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	N/A	2.539e+02	134.500	ug/g
5 SAMPLE	S98T003400 0 W	@IC4G-01 GLYCOLTI SOLID	SOLID	N/A	1.215e+02	121.600	ug/g
6 DUP	S98T003400 0 W	@IC4G-01 F*4-01 SOLID	SOLID	1.06e+02	8.28e+01	24.576	RPD
6 DUP	S98T003400 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	1.49e+03	1.34e+03	10.601	RPD
6 DUP	S98T003400 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	2.54e+02	2.42e+02	4.839	RPD
6 DUP	S98T003400 0 W	@IC4G-01 GLYCOLTI SOLID	SOLID	<1.22e2	<1.21e2		RPD
7 SPK	S98T003400 0 W	@IC4G-01 F*4-01 SOLID	SOLID	5.89e1	5.90e+01	100.170 % Recovery	
7 SPK	S98T003400 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	1.24e2	1.11e+02	89.516 % Recovery	
7 SPK	S98T003400 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	1.19e2	1.09e+02	91.597 % Recovery	
7 SPK	S98T003400 0 W	@IC4G-01 GLYCOLTI SOLID	SOLID	1.07e2	9.36e+01	87.477 % Recovery	

Final page for worklist# 27850

Analyst Signature Date

Analyst Signature Date

Reviewer Signature Date 1/14/99

LABCORE Data Entry Template for Worklist# 27850

Analyst: ADP Instrument: IC 4552 Book# 3570210
347021A
Method: LA-533-105 Rev/Mod F-0
Worklist Comment: BY-105 @IC4G-01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T003400	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
6 DUP	S98T003400	0 W	@IC4G-01	SOLID		
7 SPK	S98T003400	0 W	@IC4G-01	SOLID		

Final page for worklist # 27850

James Luge for ADP 1/18/99
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

uploaded 1-18-99 Validated 1/18/99 JM Luge
John Wauell
27850TAN.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Data Reprocessed On 01/10/1999 19:25:28

=====

Sample Name: 35N21-C LCS	Date: 01/10/1999 19:22:40
Data File : C:\DX\DATA\99011001.D02	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 2	Detector: CDM-2
Analyst : <i>Anthony Purvuta</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

1-10-98

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	25226	968182	1	
2	2.10		0.000	22803	371120	2	
3	2.25		0.000	19819	258988	2	
4	2.57		0.000	9442	58492	2	
5	2.64		0.000	8290	43882	2	
6	2.75		0.000	5038	36213	2	
8	3.13		0.000	922	17820	1	
14	4.51		0.000	2793	16817	2	
15	4.63		0.000	2932	34634	2	
16	4.87		0.000	2050	24910	2	
24	6.27		0.000	1681	19138	2	
26	7.31	fluoride	62.482106 ³¹	1843213	26833352	2	0.00
27	8.03	glycolate	106.987999	525650	9527586	2	0.00
28	8.51	acetate	114.231921 ²	592674	11434767	2	0.00
29	9.23	formate	113.928951 ¹⁴	1021310	19130883	2	0.00
30	11.31		0.000	10433892	459823017	2	
31	12.05		0.000	9463993	712396213	2	

Totals 397.629 23981729 1240996014

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 1574 TO 1583

File: 99011001.D02 Sample: 35N21-C LCS

1574

Data Reprocessed On 01/10/1999 19:46:00

```

=====
Sample Name: 34N21-A CCV                      Date: 01/10/1999 19:42:58
Data File  : C:\DX\DATA\99011001.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3             Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101  5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.40		0.000	2166	31147	2	
5	2.19		0.000	2073	25018	1	
10	2.96		0.000	3526	24564	2	
11	3.07		0.000	2977	36597	2	
13	3.44		0.000	3184	20314	2	
14	3.59		0.000	2516	40339	2	
26	5.52		0.000	2649	21163	2	
27	5.71		0.000	1997	16284	2	
28	5.91		0.000	1719	22922	2	
30	7.28	fluoride	56.060	1643515	23997934	2	0.00
31	8.00	glycolate	92.531	462715	8236192	2	0.00
32	8.48	acetate	112.785	569781	11293367	2	0.00
33	9.20	formate	116.699	1048981	19604495	2	0.00
34	11.31		0.000	10388087	453419247	2	
35	12.05		0.000	9287509	681141093	2	
Totals			378.075	23423396	1197930676		

```

=====
Sample Name: BLANK                      Date: 01/10/1999 19:02:21
Data File  : C:\DX\DATA\99011001.D01
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 1
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          1    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.97		0.000	141104	3704000	1	
2	5.89		0.000	11500	196600	1	
4	7.55	fluoride	0.081	70372	3169722	2	0.00
5	8.24	glycolate	0.054	34008	639040	2	0.00
6	9.76		0.000	490542	21227916	2	
7	10.67		0.000	12389103	174216732	2	
8	11.33		0.000	8146428	318193058	2	
9	12.03		0.000	12001123	988416101	2	

```

-----
Totals                      0.136    33284180 1509763169
-----

```

File: 99011001.D01 Sample: BLANK

```

=====
Sample Name: PREP BLANK                      Date: 01/10/1999 20:03:18
Data File  : C:\DX\DATA\99011001.D04
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 4          Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700   5Hz    0.00 19.00    16000

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	2672	42584	1	
7	2.59		0.000	1769	17300	1	
15	4.21		0.000	1786	23446	2	
16	4.53		0.000	1410	30621	2	
20	5.48		0.000	2542	27230	2	
21	5.67		0.000	1937	21242	2	
23	6.27		0.000	1978	43842	2	
25	7.01		0.000	2174	17213	2	
26	7.17		0.000	3628	67150	2	
27	7.63	fluoride	0.009	3511	42649	2	0.00
28	7.92		0.000	3461	22195	2	
29	8.08	glycolate	-0.010	2455	85310	2	0.00
31	9.07		0.000	2238	16874	2	
32	9.25	formate	0.005	2791	28721	2	0.00
35	11.31		0.000	10460441	458896946	2	
36	12.03		0.000	9028288	664707599	2	
Totals			0.004	19523081	1124090922		

File: 99011001.D04 Sample: PREP BLANK

Data Reprocessed On 01/14/1999 11:46:11

```

=====
Sample Name: S98T003400 SAM                      Date: 01/10/1999 20:24:07
Data File  : E:\DATA\99011001.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          11    5700   5Hz   6.99 10.66    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	5981	83772	1	
4	1.70		0.000	10742	126371	2	
5	2.10		0.000	4350	65108	1	
6	2.83		0.000	8760	133680	2	
7	3.13		0.000	9059	207792	2	
8	3.57		0.000	8049	114260	2	
9	3.92		0.000	5304	47572	2	
10	4.03		0.000	4145	31917	2	
11	4.17		0.000	2962	48750	2	
12	4.57		0.000	4437	24344	2	
13	4.64		0.000	4323	24419	2	
22	5.83		0.000	890	31616	1	
24	6.64		0.000	4449	29081	2	
25	7.25	fluoride	0.535	107117	1755445	2	0.00
26	8.45	acetate	7.529	344414	7017876	2	0.00
27	9.15	formate	1.287	99771	1914769	2	0.00
28	9.89		0.000	9325	146366	1	
29	10.88		0.000	5639007	57927278	2	
30	11.17		0.000	48140320	527609083	2	
31	11.65		0.000	12952105	134791142	2	
32	11.92		0.000	65454207	701283499	2	
33	12.24		0.000	22388731	703971881	2	
Totals			9.350	155208447	2137386022		

Data Reprocessed On 01/14/1999 11:47:13

```

=====
Sample Name: S98T003400 DUP                      Date: 01/10/1999 20:44:16
Data File  : E:\DATA\99011001.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6                Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          11    5700   5Hz   6.24 10.66      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	6907	186280	1	
3	1.73		0.000	22897	208400	1	
4	2.10		0.000	3099	41152	1	
5	2.68		0.000	9017	126958	2	
6	2.93		0.000	13134	108912	2	
7	3.08		0.000	10372	139004	2	
8	3.43		0.000	3868	19750	2	
9	3.60		0.000	4046	27290	2	
10	3.69		0.000	2685	30020	2	
16	4.63		0.000	3845	50128	2	
21	5.55		0.000	2977	57975	2	
22	5.87		0.000	4417	88247	2	
23	6.64		0.000	3323	41718	2	
24	7.25	fluoride	0.420	71724	1302999	2	0.00
25	8.48	acetate	6.777	307716	6341001	2	0.00
26	9.17	formate	1.226	94367	1822751	2	0.00
27	9.97		0.000	5891	84308	1	
28	10.88		0.000	4861560	54067281	2	
29	11.17		0.000	42482573	489850594	2	
30	11.65		0.000	12593865	130759627	2	
31	11.97		0.000	72768060	809536831	2	
32	12.21		0.000	43313707	919861208	2	
Totals			8.423	176590052	2414752434		

Data Reprocessed On 01/14/1999 11:48:52

Sample Name: S98T003400	SPK	Date: 01/10/1999 21:13:02
Data File :	E:\DATA\99011001.D07	
Method :	C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1	System: 2	Inject#: 7
Analyst :	Column: AG14A-SC,AS14A-SC,	SRS
		Detector:CDM-2

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	11	5700	5Hz	6.25	10.50	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	3766	54237	2	
3	1.13		0.000	1649	16413	2	
4	1.70		0.000	16000	170064	2	
8	3.29		0.000	1505	16058	2	
9	3.73		0.000	2350	31123	2	
10	3.88		0.000	2332	16251	2	
11	4.13		0.000	2878	40503	2	
12	4.43		0.000	3482	54419	2	
13	4.71		0.000	3820	35107	2	
14	4.97		0.000	4972	95174	2	
15	5.28		0.000	5161	102086	2	
16	5.61		0.000	5481	51434	2	
17	5.76		0.000	5403	43153	2	
18	5.93		0.000	5376	60076	2	
19	6.11		0.000	4209	30980	2	
20	7.17	fluoride	6.434	1798982	25327305	2	0.00
21	7.87	glycolate	9.361	463829	7651057	2	0.00
22	8.35	acetate	18.632	904423	16982714	2	0.00
23	9.01	formate	12.209	1057726	18818933	3	0.00
24	9.76		0.000	5270	50960	4	
25	10.88		0.000	9199834	132107616	2	
26	11.17		0.000	48059395	566901304	2	
27	11.68		0.000	14314727	140650256	2	
28	11.92		0.000	65984095	727220394	2	
29	12.24		0.000	23043253	709800185	2	
Totals			46.636	164899916	2346327802		

LABCORE Completed Worklist Report for Worklist# 27859

Analyst: adp Instrument: IC45S2 Book#: 35N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 RERUN D.F.=6 JHW

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.05e1	6.42e+01	106.116 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.19e2	9.76e+01	82.017 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.28e2	1.21e+02	94.531 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.16e2	1.12e+02	96.552 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.21e1	5.86e+01	94.364 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.23e2	1.15e+02	93.496 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.30e2	1.22e+02	93.846 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.02e2	9.76e+01	95.686 % Recovery	
4 BLNK-PREP	0	@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002994	0 W @IC4G-01 F*4-01	SOLID	N/A	1.728e+03	36.320 ug/g	
5 SAMPLE	S98T002994	0 W @IC4G-01 ACETATE2	SOLID	N/A	4.899e+02	23.430 ug/g	
5 SAMPLE	S98T002994	0 W @IC4G-01 FORMATE2	SOLID	N/A	2.644e+02	72.660 ug/g	
6 DUP	S98T002994	0 W @IC4G-01 F*4-01	SOLID	1.73e+03	2.10e+03	19.321 RPD	
6 DUP	S98T002994	0 W @IC4G-01 ACETATE2	SOLID	4.90e+02	3.71e+02	27.642 RPD	
6 DUP	S98T002994	0 W @IC4G-01 FORMATE2	SOLID	2.64e+02	2.35e+02	11.623 RPD	
6 DUP	S98T002994	0 W @IC4G-01 GLYCOLT1	SOLID	?	<6.60e1	RPD	
7 SPK	S98T002994	0 W @IC4G-01 F*4-01	SOLID	6.05e1	7.56e+01	124.959 % Recovery	
7 SPK	S98T002994	0 W @IC4G-01 ACETATE2	SOLID	1.19e2	9.92e+01	83.361 % Recovery	
7 SPK	S98T002994	0 W @IC4G-01 FORMATE2	SOLID	1.28e2	1.26e+02	98.438 % Recovery	
7 SPK	S98T002994	0 W @IC4G-01 GLYCOLT1	SOLID	1.16e2	0.00e+00	0.000 % Recovery	
8 SAMPLE	S98T002995	0 W @IC4G-01 F*4-01	SOLID	N/A	2.897e+03	36.670 ug/g	
8 SAMPLE	S98T002995	0 W @IC4G-01 ACETATE2	SOLID	N/A	4.410e+02	23.660 ug/g	
9 DUP	S98T002995	0 W @IC4G-01 F*4-01	SOLID	2.90e+03	6.61e+03	78.023 RPD	
9 DUP	S98T002995	0 W @IC4G-01 ACETATE2	SOLID	4.41e+02	5.61e+02	23.952 RPD	
9 DUP	S98T002995	0 W @IC4G-01 FORMATE2	SOLID	?	3.15e+02	RPD	
9 DUP	S98T002995	0 W @IC4G-01 GLYCOLT1	SOLID	?	<6.65e1	RPD	
10 SPK	S98T002995	0 W @IC4G-01 F*4-01	SOLID	6.05e1	6.83e+01	112.893 % Recovery	
10 SPK	S98T002995	0 W @IC4G-01 ACETATE2	SOLID	1.19e2	1.02e+02	85.714 % Recovery	
10 SPK	S98T002995	0 W @IC4G-01 FORMATE2	SOLID	1.28e2	0.00e+00	0.000 % Recovery	
10 SPK	S98T002995	0 W @IC4G-01 GLYCOLT1	SOLID	1.16e2	0.00e+00	0.000 % Recovery	
11 SAMPLE	S98T002996	0 W @IC4G-01 F*4-01	SOLID	N/A	2.487e+03	36.820 ug/g	
12 DUP	S98T002996	0 W @IC4G-01 F*4-01	SOLID	2.49e+03	3.17e+03	24.028 RPD	
12 DUP	S98T002996	0 W @IC4G-01 ACETATE2	SOLID	?	1.91e+02	RPD	
12 DUP	S98T002996	0 W @IC4G-01 FORMATE2	SOLID	?	1.42e+02	RPD	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27859

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
12	DUP	S98T002996	0	W	@IC4G-01 GLYCOLTI	SOLID	?	<6.59e1		RPD

Final page for worklist# 27859

_____
Analyst Signature   Date

_____
Analyst Signature   Date

1/26/99
_____
Reviewer Signature   Date

01/12/99 11:49
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27859

Analyst: ADP Instrument: IC _____, Book# 35N21-A LCS
Method: LA-533-105 Rev/Mod F-0 34N21-A CCV
Worklist Comment: BY105 RERUN D.F.=6 JHW

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002994	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01, FORMATE2					
6 DUP	S98T002994	0 W	@IC4G-01	SOLID		
7 SPK	S98T002994	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T002995	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01					
9 DUP	S98T002995	0 W	@IC4G-01	SOLID		
10 SPK	S98T002995	0 W	@IC4G-01	SOLID		
11 SAMPLE	S98T002996	0 W	@IC4G-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: F*4-01					
12 DUP	S98T002996	0 W	@IC4G-01	SOLID		

Final page for worklist # 27859

Analyst Signature Date 01-21-99

_____
Analyst Signature Date

Data Entry Comments:

up loaded 1-21-99John W. Hall
27859JAN.CSVvalidated 1/26/99 gm faye
RPD's still high

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 35N21-A LCS Date: 01/21/1999 02:42:01
 Data File : C:\DX\DATA\99012001.D16
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: System: 2, Inject#: 16 Detector: CDM-2
 Analyst : *Anthony Parnis* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 5700 5Hz 0.00 19.00 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
10	3.84		0.000	2982	72204	2	
20	6.00		0.000	3403	75604	2	
21	7.15	fluoride	64.211106 ¹³	1857350	27598448	2	0.00
22	7.84	glycolate	111.72096 ³¹	543705	9952243	2	0.00
23	8.32	acetate	97.58382 ⁰⁰	500250	9806006	2	0.00
24	9.07	formate	121.11394 ⁶²	1074732	20359843	2	0.00
Totals			394.626	3982422	67864349		

File: 99012001.D16 Sample: 35N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1587 TO 1560.

=====

Sample Name: 34N21-A CCV	Date: 01/21/1999 03:02:59
Data File : C:\DX\DATA\99012001.D17	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 17	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
32	7.17	fluoride	58.647	1702108	25138984	2	0.00
33	7.87	glycolate	97.606	471861	8688568	2	0.00
34	8.35	acetate	114.891	597205	11499268	2	0.00
35	9.09	formate	122.403	1095485	20580827	2	0.00
Totals			393.547	3866660	65907647		

File: 99012001.D17 Sample: 34N21-A CCV

```

=====
Sample Name: BLANK                      Date: 01/21/1999 02:21:49
Data File  : C:\DX\DATA\99012001.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 15      Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	1	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	16016	225463	2	
2	0.47		0.000	11811	132090	2	
3	1.97		0.000	17826	140797	2	
4	2.23		0.000	4315	98813	2	
8	2.87		0.000	3990	57287	2	
11	3.76		0.000	4666	77685	2	
20	5.83		0.000	5894	173005	2	
21	6.77		0.000	1399	122035	1	
Totals			0.000	65918	1027174		

File: 99012001.D15 Sample: BLANK

```
=====
Sample Name: PREP BLANK                      Date: 01/21/1999 03:23:23
Data File  : C:\DX\DATA\99012001.D18
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 18         Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====
```

```
=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1  5700  5Hz   0.00 19.00   50000
=====
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.27		0.000	3144	53730	1	
23	5.60		0.000	3323	60798	2	
35	8.88		0.000	3151	93218	2	
Totals			0.000	9619	207746		

File: 99012001.D18 Sample: PREP BLANK

Data Reprocessed On 01/21/1999 14:03:30

```

=====
Sample Name: S98T002994 SAM                      Date: 01/21/1999 03:46:07
Data File  : E:\DATA\99012001.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6 5700 5Hz 0.00 19.00 50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.70		0.000	102523	1187238	1	
5	2.77		0.000	7591	128689	3	
21	5.71		0.000	6482	119301	2	
23	7.07	fluoride	8.850	4221995	66039533	2	0.00
24	8.21	acetate	2.509	199164	4378665	2	0.00
25	8.93	formate	1.354	191287	3742940	2	0.00
26	9.73		0.000	30462	564606	2	
Totals			12.713	4759504	76160972		

File: 99012001.D19 Sample: S98T002994 SAM

```

=====
Sample Name: S98T002994  DUP                      Date: 01/21/1999 04:06:21
Data File  : C:\DX\DATA\99012001.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      6    5700    5Hz    0.00  19.00      50000
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	105326	1323298	1	
4	2.85		0.000	15376	242029	2	
5	2.93		0.000	15395	85260	2	
6	3.11		0.000	15593	135193	2	
7	3.20		0.000	15572	284026	2	
8	3.57		0.000	11268	71300	2	
9	3.69		0.000	12027	93645	2	
10	3.87		0.000	12745	97669	2	
11	4.04		0.000	13543	348932	2	
12	4.60		0.000	9077	58929	2	
13	4.68		0.000	8916	84415	2	
15	4.97		0.000	10813	214284	2	
16	5.39		0.000	8209	85644	2	
17	5.63		0.000	7146	60416	2	
18	5.73		0.000	8285	99446	2	
20	7.09	fluoride	10.715	5122819	80766389	2	0.00
21	8.24	acetate	1.889	149384	3352157	2	0.00
22	8.96	formate	1.197	164811	3302503	2	0.00
23	9.76		0.000	32098	622169	2	
Totals			13.800	5738403	91327704		

File: 99012001.D20 Sample: S98T002994 DUP

Minutes
1592

Data Reprocessed On 01/21/1999 14:04:23

```

=====
Sample Name: S98T002994  SPK                      Date: 01/21/1999 04:26:23
Data File   : E:\DATA\99012001.D21
Method      : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 21             Detector: CDM-2
Analyst     :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5700  5Hz   0.00 19.00      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	102660	1158268	2	
5	2.83		0.000	15244	236632	2	
6	2.96		0.000	14170	205206	2	
7	3.24		0.000	13468	77309	2	
8	3.32		0.000	13056	83617	2	
9	3.48		0.000	15031	366527	2	
10	3.89		0.000	12656	117746	2	
11	4.08		0.000	10441	90976	2	
12	4.23		0.000	10412	67165	2	
13	4.39		0.000	11228	120540	2	
14	4.55		0.000	12043	157164	2	
15	4.79		0.000	11307	124921	2	
16	4.99		0.000	10997	124194	2	
17	5.23		0.000	9381	61765	2	
18	5.36		0.000	9547	53846	2	
19	5.45		0.000	9680	116662	2	
20	5.67		0.000	9794	52526	2	
21	5.76		0.000	9607	55020	2	
22	5.89		0.000	9113	128517	2	
23	7.07	fluoride	12.629	6064151	96194357	2	0.00
24	7.73	glycolate	5.893	459609	8830521	2	0.00
25	8.21	acetate	7.467	634042	12555184	2	0.00
26	8.93	formate	7.675	1119393	21746035	2	0.00
27	9.73		0.000	31232	593397	2	
Totals			33.665	8618260	143318095		

Data Reprocessed On 01/21/1999 14:06:09

Sample Name: S98T002995	Date: 01/21/1999 09:33:52
Data File : E:\DATA\99012101.D25	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 25	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	4739	5Hz	0.00	15.79	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	5001	54864	1	
2	1.70		0.000	96643	1126544	1	
3	2.68		0.000	12671	99717	2	
4	2.80		0.000	12874	79871	2	
5	2.95		0.000	13254	233747	2	
6	3.20		0.000	11979	85290	2	
7	3.33		0.000	11001	196975	2	
15	4.63		0.000	4437	66374	2	
17	5.13		0.000	5456	83347	2	
18	5.41		0.000	8908	304583	2	
19	7.07	fluoride	14.692	7143011	113183295	3	0.00
20	7.73	glycolate	0.042	22751	229523	4	0.00
21	8.21	acetate	2.237	186403	3927662	2	0.00
22	8.93	formate	1.438	203647	3979483	2	0.00
23	9.73		0.000	34653	641681	2	
Totals			18.409	7772688	124292956		

File: 99012101.D25 Sample: S98T002995

Data Reprocessed On 01/21/1999 14:06:56

```

=====
Sample Name: S98T002995 DUP                      Date: 01/21/1999 09:51:22
Data File  : E:\DATA\99012101.D26
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 26              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6  4947  5Hz  0.00 16.49   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.70		0.000	116294	1363712	1	
3	2.81		0.000	14125	195307	2	
4	2.93		0.000	13983	101059	2	
5	3.07		0.000	12815	196435	3	
14	4.87		0.000	4649	62412	2	
20	7.07	fluoride	33.431	17277713	291127987	2	0.00
21	8.19	acetate	2.837	228823	4920568	2	0.00
22	8.91	formate	1.590	219816	4405768	2	0.00
23	9.68		0.000	31007	650910	2	

```

-----
Totals                      37.857  17919226  303024158
=====

```

File: 99012101.D26 Sample: S98T002995 DUP

Data Reprocessed On 01/21/1999 14:07:43

```

=====
Sample Name: S98T002995 SPK                      Date: 01/21/1999 10:06:54
Data File  : E:\DATA\99012101.D27
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 27              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1           6  4434 5Hz  0.00 14.78      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.70		0.000	101441	1205568	1	
3	2.85		0.000	16278	241119	2	
4	2.93		0.000	16665	272886	2	
5	3.23		0.000	13762	138347	2	
6	3.56		0.000	9303	191686	2	
7	3.88		0.000	7009	103639	2	
12	5.13		0.000	2867	58305	2	
14	5.64		0.000	4183	70961	2	
15	7.07	fluoride	18.106	8850263	142207790	2	0.00
16	7.73	glycolate	5.853	468864	8769992	2	0.00
17	8.21	acetate	7.333	617768	12334027	2	0.00
18	8.93	formate	7.602	1130225	21535708	2	0.00
19	9.73		0.000	36436	726910	2	
Totals			38.894	11275063	187856936		

File: 99012101.D27 Sample: S98T002995 SPK

```

=====
Sample Name: S98T002996                      Date: 01/21/1999 13:00:50
Data File  : C:\DX\DATA\99012101.D28
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 28          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6  5700  5Hz   0.00 19.00      50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	5434	93592	2	
2	1.70		0.000	266888	3524470	2	
3	2.52		0.000	15468	884000	1	
9	5.73		0.000	3419	64580	1	
12	7.12	fluoride	12.561	6072025	95633042	2	0.00
13	8.24	acetate	1.031	91053	1932038	2	0.00
14	8.96	formate	0.682	97582	1858951	2	0.00
15	9.71		0.000	13205	201480	1	
Totals			14.273	6565073	104192153		

File: 99012101.D28 Sample: S98T002996

Data Reprocessed On 01/21/1999 14:08:56

```

=====
Sample Name: S98T002996 DUP                      Date: 01/21/1999 13:20:20
Data File  : E:\DATA\99012101.D29
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 29              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6   5560  5Hz   0.00 18.53   50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.37		0.000	12690	273426	2	
3	1.70		0.000	98260	1203312	1	
11	3.77		0.000	6523	58333	2	
14	4.19		0.000	3739	53495	2	
19	4.97		0.000	5153	65753	2	
30	7.09	fluoride	16.140	7867732	125347769	2	0.00
31	8.21	acetate	0.975	86598	1838884	2	0.00
32	8.93	formate	0.726	101097	1982509	2	0.00
33	9.73		0.000	15138	256924	1	

Totals ----- 17.841 8196931 131080405

File: 99012101.D29 Sample: S98T002996 DUP

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 27885

Analyst: ehc

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-O

Worklist Comment: BY-105 for @IC4G-01

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC		1	2.40e-02	0.024	ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC		1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC		1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC		1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC		6.10e1	6.41e+01	105.082 %	Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC		1.16e2	1.16e+02	100.000 %	Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC		1.19e2	1.28e+02	107.563 %	Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLTI	QC		1.08e2	1.09e+02	100.926 %	Recovery
3 CCV	0	@IC4G-QC F*4	QC		6.29e1	5.71e+01	90.779 %	Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC		1.20e2	1.19e+02	99.167 %	Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC		1.20e2	1.23e+02	102.500 %	Recovery
3 CCV	0	@IC4G-QC GLYCOLTI	QC		1.03e2	9.68e+01	93.981 %	Recovery
4 SAMPLE	S98T002814	0	@IC4G-01 F*4-01	LIQUID	N/A	1.202e+03	34.440	ug/mL
4 SAMPLE	S98T002814	0	@IC4G-01 ACETATE2	LIQUID	N/A	3.625e+03	64.440	ug/mL
4 SAMPLE	S98T002814	0	@IC4G-01 FORMATE2	LIQUID	N/A	3.037e+03	68.880	ug/mL
4 SAMPLE	S98T002814	0	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 6.222e+01	62.220	ug/mL
5 DUP	S98T002814	0	@IC4G-01 F*4-01	LIQUID	1.20e+03	4.19e+03	0.837	RPD
5 DUP	S98T002814	0	@IC4G-01 ACETATE2	LIQUID	3.62e+03	3.62e+03	0.000	RPD
5 DUP	S98T002814	0	@IC4G-01 FORMATE2	LIQUID	3.04e+03	3.05e+03	0.328	RPD
5 DUP	S98T002814	0	@IC4G-01 GLYCOLTI	LIQUID	<6.22e1	<6.22e1		RPD
6 SPK	S98T002814	0	@IC4G-01 F*4-01	LIQUID	6.10e1	2.55e+01	41.803 %	Recovery
6 SPK	S98T002814	0	@IC4G-01 ACETATE2	LIQUID	1.16e2	3.90e+01	33.621 %	Recovery
6 SPK	S98T002814	0	@IC4G-01 FORMATE2	LIQUID	1.19e2	4.79e+01	40.252 %	Recovery
6 SPK	S98T002814	0	@IC4G-01 GLYCOLTI	LIQUID	1.08e2	1.14e+02	105.556 %	Recovery

Final page for worklist# 27885

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature   Date 3/18/99

LABCORE Data Entry Template for Worklist# 27885

Analyst: CHC Instrument: IC 4552 Book# 59N21-A
Method: LA-533-105 Rev/Mod F-0 58N21-A

Worklist Comment: BY-105 for @IC4G-01 Try 60% DF to 1111DF

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	LCS-INST				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	SAMPLE	S98T002814	0		@IC4G-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested:				ACETATE2, F*4-01	FORMATE2, GLYCOLT1		
5	DUP	S98T002814	0		@IC4G-01	LIQUID		
6	SPK	S98T002814	0		@IC4G-01	LIQUID		

Final page for worklist # 27885

CHC 3-16-99
Signature Date

Signature Date

Data Entry Comments:

uploaded 3-16-99

John Howell

27885MAR.CSV

validated 3/17/99 JMT

There appear to be matrix

interferences with spk on this

sample - We will rerun if

1601

desired. JMT

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

Sample Name: INSTR BLAMK Date: 03/16/1999 08:24:29
 Data File : C:\DX\DATA\99031601.D10
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 10 Detector: CDM-2
 Analyst : *Eld Ebr* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 4726 5Hz 0.00 15.75 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	13810	142468	1	
2	2.13		0.000	26421	404880	2	
3	2.40		0.000	131183	5063043	2	
4	4.72		0.000	1944703	178458176	2	
5	7.15	fluoride	0.024	4000	60400	1	0.00
Totals			0.024	2120117	184128966		

File: 99031601.D10 Sample: INSTR BLAMK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1602 TO 1609.

Sample Name: 59N21-A LCS Date: 03/16/1999 09:16:35
 Data File : C:\DX\DATA\99031601.D13
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 13 Detector: CDM-2
 Analyst : *P.H. Lohm* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4593	5Hz	0.00	15.31	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	3637	39332	1	
2	2.23		0.000	4257	66709	2	
3	2.52		0.000	2458	19517	2	
4	2.64		0.000	3217	28542	2	
5	2.85		0.000	4513	25831	2	
6	2.93		0.000	5038	76960	2	
8	3.43		0.000	2604	30225	2	
10	3.85		0.000	1875	21298	2	
14	5.11		0.000	1930	20404	2	
17	5.65		0.000	1862	18483	2	
25	7.15	fluoride	64.138	2005278	29313765	2	0.00
26	7.84	glycolate	108.701	543876	9883028	2	0.00
27	8.32	acetate	115.774	554574	10588035	2	0.00
28	9.01	formate	128.190	1042471	19742134	2	0.00
Totals			416.803	4177592	69874264		

File: 99031601.D13 Sample: 59N21-A LCS

=====

Sample Name: 58N21-A CCV	Date: 03/16/1999 09:00:28
Data File : C:\DX\DATA\99031601.D12	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 12	Detector: CDM-2
Analyst : <i>QNP/Chv</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4520	5Hz	0.00	15.06	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	11133	495920	1	
2	2.19		0.000	10822	244028	2	
3	2.29		0.000	6426	60407	2	
6	2.84		0.000	3421	33466	2	
8	3.09		0.000	3804	28070	2	
10	3.37		0.000	1949	19663	2	
16	4.39		0.000	899	24264	2	
22	5.63		0.000	2888	21171	2	
23	5.79		0.000	2168	18850	2	
28	7.20	fluoride	57.105	1774393	25943485	2	0.00
29	7.89	glycolate	96.847	488529	8762490	2	0.00
30	8.37	acetate	118.621	559200	10852361	2	0.00
31	9.07	formate	123.485	1018336	18981238	2	0.00
Totals			396.058	3883967	65485413		

File: 99031601.D12 Sample: 58N21-A CCV

Data Reprocessed On 03/16/1999 14:48:49

```

=====
Sample Name: S98T002814                      Date: 03/16/1999 09:34:43
Data File  : E:\DATA\99031601.D14
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1111  4566  5Hz   6.39 10.86      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	8394	274284	1	
2	1.70		0.000	222656	2770564	1	
3	3.09		0.000	16658	478644	2	
4	3.25		0.000	11367	206225	3	
7	3.77		0.000	2174	24641	2	
8	4.01		0.000	3393	18144	2	
9	4.16		0.000	5716	104102	2	
10	4.43		0.000	6199	25757	2	
11	4.55		0.000	7479	63564	2	
12	4.65		0.000	8096	46158	2	
13	4.76		0.000	8793	42952	2	
14	4.89		0.000	10374	182785	2	
15	5.17		0.000	11639	94048	2	
16	5.37		0.000	13131	519110	2	
17	6.96	fluoride	1202.469	2991259	51189926	2	0.00
18	8.11	acetate	3625.326	1265945	30189278	2	0.00
19	8.80	formate	3036.834	2192639	44159387	2	0.00
20	9.55		0.000	356856	7485623	3	
21	10.11		0.000	6023	47496	4	
Totals			7864.630	7148791	137922687		

Data Reprocessed On 03/16/1999 14:50:48

```

=====
| Sample Name: S98T002814 DUP                      Date: 03/16/1999 09:49:58
| Data File  : E:\DATA\99031601.D15
| Method     : C:\DX\METHOD\MAS14GLY.MET
| ACI Address: 1 System: 2 Inject#: 15              Detector: CDM-2
| Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1111 4121 5Hz 6.28 10.27 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	4215	64308	1	
2	1.70		0.000	221140	2649416	1	
3	3.04		0.000	5421	255185	2	
5	3.19		0.000	757	24995	2	
8	3.93		0.000	2942	16231	2	
9	4.03		0.000	3763	24410	2	
10	4.19		0.000	4433	327201	2	
11	4.99		0.000	9721	120723	2	
12	5.19		0.000	11428	578005	2	
13	6.93	fluoride	1194.222	2945747	50823616	2	0.00
14	8.11	acetate	3623.738	1278852	30176210	2	0.00
15	8.77	formate	3053.438	2162708	44413831	2	0.00
16	9.55		0.000	364600	7604294	3	
17	10.08		0.000	7147	72131	4	
Totals			7871.398	7022874	137150556		

```

=====
Sample Name: S98T002814 SPK                      Date: 03/16/1999 11:41:10
Data File  : C:\DX\DATA\99031601.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1111 4079 5Hz 0.00 13.59 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	2841	31642	1	
2	1.70		0.000	251901	2323928	1	
3	3.00		0.000	17228	432010	2	
4	3.21		0.000	15133	100218	2	
5	3.33		0.000	14710	294996	2	
6	3.68		0.000	10656	167510	2	
7	4.00		0.000	6933	38346	2	
8	4.16		0.000	5882	55814	2	
9	4.29		0.000	3583	32479	3	
14	5.49		0.000	1436	202431	1	
15	7.09	fluoride	1459.801	3966232	62675976	2	0.00
16	7.76	glycolate	1149.090	539777	9481664	2	0.00
17	8.21	acetate	4018.921	1547748	33421700	2	0.00
18	8.91	formate	3520.248	2732114	51612858	2	0.00
19	9.63		0.000	283103	5504830	2	
Totals			10148.060	9399278	166376403		

File: 99031601.D16 Sample: S98T002814 SPK

LABCORE Completed Worklist Report for Worklist# 27933

Analyst: adp

Instrument: IC45S2

Book#: 58N21A

Method: LA-533-105 Rev/Mod F-6

Worklist Comment: RERUN BY105 DF=2121 MAY BE HIGH RPD

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	6.09e+01	99.836 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.12e+02	96.552 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.21e+02	101.681 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.08e2	1.02e+02	94.444 % Recovery	
3 CCV	0	@IC4G-QC F*4	QC	6.29e1	5.83e+01	92.687 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.18e+02	98.333 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.24e+02	103.333 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.03e2	9.78e+01	94.951 % Recovery	
4 SAMPLE	S98T003006	@IC4G-01 F*4-01	LIQUID	N/A	2.478e+02	65.750 ug/mL	
4 SAMPLE	S98T003006	@IC4G-01 ACETATE2	LIQUID	N/A	1.955e+03	123.000 ug/mL	
4 SAMPLE	S98T003006	@IC4G-01 FORMATE2	LIQUID	N/A	1.281e+03	131.500 ug/mL	
4 SAMPLE	S98T003006	@IC4G-01 GLYCOLT1	LIQUID	N/A	< 1.188e+02	118.800 ug/mL	
5 DUP	S98T003006	@IC4G-01 F*4-01	LIQUID	2.48e+02	2.51e+02	1.202 RPD	
5 DUP	S98T003006	@IC4G-01 ACETATE2	LIQUID	1.96e+03	1.99e+03	1.519 RPD	
5 DUP	S98T003006	@IC4G-01 FORMATE2	LIQUID	1.28e+03	1.30e+03	1.550 RPD	
5 DUP	S98T003006	@IC4G-01 GLYCOLT1	LIQUID	<1.19e2	<1.19e2	RPD	

Final page for worklist# 27933

Analyst Signature

Date

Analyst Signature

Date

Janet M. Lyle
Reviewer Signature Date 2/25/99

LABCORE Data Entry Template for Worklist# 27933

Analyst: ADD Instrument: IC 4552 Book# 59N21-A LCS
Method: LA-533-105 Rev/Mod F-0 58N21-A CCV

Worklist Comment: RERUN BY105 DF=2121 MAY BE HIGH RPD

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	LCS-INST				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	SAMPLE	S98T003006 0			@IC4G-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
5	DUP	S98T003006 0			@IC4G-01	LIQUID		

Final page for worklist # 27933

Anthony P. Wansell 02-13-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 2-17-99
John Wansell

Validated 2/24/99 gmtuf
Fated 2-13-99

27933FEB.C6V

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: BLANK Date: 02/13/1999 07:11:43
 Data File : E:\DATA\99021301.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst : *MF/ADP* Column: AG14A-SC, AS14A-SC, SRS
3/5/99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	21676	274340	1	
2	2.19		0.000	33083	413339	2	
3	2.45		0.000	95877	1572519	2	
4	2.97		0.000	311285	22604638	2	
5	4.21		0.000	322197	15856056	2	
6	4.97		0.000	441040	5446455	2	
7	5.21		0.000	485875	5531456	2	
8	5.36		0.000	513452	17146572	2	
9	7.92	glycolate	11.666	2003355	198410020	2	0.00
Totals			11.666	4227840	267255395		

File: 99021301.D01 Sample: BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1612 TO 1616.

Data Reprocessed On 02/17/1999 11:38:22

Sample Name: 59N21-A LCS	Date: 02/13/1999 07:33:36
Data File : E:\DATA\99021301.D02	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 2	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	35002	1221060	1	
2	2.10		0.000	27573	445969	2	
3	2.29		0.000	27263	402874	2	
4	2.55		0.000	22056	250520	2	
5	2.77		0.000	20223	198317	2	
6	2.96		0.000	17402	98373	2	
7	3.04		0.000	16919	104396	2	
8	3.16		0.000	13924	77008	2	
9	3.31		0.000	12935	243675	2	
17	4.95		0.000	2044	16569	2	
18	5.11		0.000	2819	25967	2	
19	5.37		0.000	3495	19868	2	
23	5.92		0.000	4230	61724	2	
25	6.75		0.000	4468	36107	2	
26	7.28	fluoride	60.902	1888927	27762242	2	0.00
27	7.97	glycolate	102.112	503857	9259406	2	0.00
28	8.48	acetate	111.845	524695	10223047	2	0.00
29	9.20	formate	120.651	982096	18523446	3	0.00
Totals			395.511	4109928	68970570		

File: 99021301.D02 Sample: 59N21-A LCS

1613

Minutes

```

=====
Sample Name: 58N21-A CCV                               Date: 02/13/1999 07:59:34
Data File  : C:\DX\DATA\99021301.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3                     Detector: CDM-2
Analyst    :                                           Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz   0.00 19.00   16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.53		0.000	18661	734508	1	
2	2.29		0.000	2042	21364	1	
10	3.67		0.000	2022	17589	2	
11	3.80		0.000	2602	19411	2	
15	4.49		0.000	3588	28998	2	
27	7.28	fluoride	58.304	1826404	26517591	2	0.00
28	7.97	glycolate	97.829	489591	8855084	2	0.00
29	8.45	acetate	118.416	554666	10833291	2	0.00
30	9.17	formate	124.032	1017179	19069631	2	0.00
Totals			398.581	3916754	66097466		

File: 99021301.D03 Sample: 58N21-A CCV

```

=====
Sample Name: S98T003006 SAM                      Date: 02/13/1999 11:14:51
Data File  : C:\DX\DATA\99021301.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9                Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          2121   5700  5Hz   0.00 19.00   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	4892	125880	1	
3	1.73		0.000	42641	428802	1	
4	2.97		0.000	10837	262778	2	
5	3.08		0.000	10764	75460	2	
6	3.21		0.000	9522	73675	2	
7	3.41		0.000	7980	65768	2	
8	3.51		0.000	7831	80013	2	
9	3.93		0.000	5298	122551	3	
16	5.00		0.000	902	16083	2	
17	5.23		0.000	3344	43654	2	
18	5.44		0.000	4619	57554	2	
19	5.68		0.000	6968	240568	2	
20	7.23	fluoride	247.804	277768	4450606	2	0.00
21	8.40	acetate	1954.517	398606	8476901	2	0.00
22	9.12	formate	1281.298	480898	8966156	2	0.00
23	9.89		0.000	75337	1372916	1	
Totals			3483.619	1348208	24859366		

File: 99021301.D09 Sample: S98T003006 SAM

```

=====
Sample Name: S98T003006 DUP                      Date: 02/13/1999 11:35:39
Data File  : C:\DX\DATA\99021301.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 2121 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	4987	196196	1	
2	1.73		0.000	41248	400934	1	
3	3.01		0.000	6515	182992	2	
4	3.09		0.000	6132	54249	3	
12	5.09		0.000	3558	56062	2	
13	5.41		0.000	5504	65312	2	
14	5.57		0.000	7636	36104	2	
15	5.68		0.000	8165	288074	2	
16	7.23	fluoride	250.712	281754	4515723	2	0.00
17	8.40	acetate	1993.574	408250	8650076	2	0.00
18	9.12	formate	1303.010	491577	9130604	2	0.00
19	9.89		0.000	77167	1394972	1	
Totals			3547.295	1342492	24971298		

File: 99021301.D10 Sample: S98T003006 DUP

LABCORE Completed Worklist Report for Worklist# 27935

Analyst: adp Instrument: IC45S2 Book#: 59N21A

Method: LA-533-105 Rev/Mod E-O

Worklist Comment: BY-105 RERUN DF=6

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	2.70e-02	0.027	ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	5.00e-02	0.050	ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	6.30e+01	103.279	% Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.16e+02	100.000	% Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.13e2	1.24e+02	104.202	% Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLTI	QC	1.08e2	1.06e+02	98.148	% Recovery
3 CCV	0	@IC4G-QC F*4	QC	6.23e1	5.79e+01	92.051	% Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.22e+02	101.667	% Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.22e+02	101.667	% Recovery
3 CCV	0	@IC4G-QC GLYCOLTI	QC	1.03e2	9.63e+01	93.495	% Recovery
4 BLNK-PREP	0	@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2	SOLID	1	2.80e-02	0.028	ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002934	0 W @IC4G-01 ACETATE2	SOLID	N/A	4.203e+02	23.450	ug/g
6 DUP	S98T002934	0 W @IC4G-01 F*4-01	SOLID	?	7.09e+02		RPD
6 DUP	S98T002934	0 W @IC4G-01 ACETATE2	SOLID	4.20e+02	3.82e+02	9.476	RPD
6 DUP	S98T002934	0 W @IC4G-01 FORMATE2	SOLID	?	2.27e+02		RPD
6 DUP	S98T002934	0 W @IC4G-01 GLYCOLTI	SOLID	?	<6.61e1		RPD
7 SAMPLE	S98T002935	0 W @IC4G-01 F*4-01	SOLID	N/A	7.770e+02	37.500	ug/g
7 SAMPLE	S98T002935	0 W @IC4G-01 ACETATE2	SOLID	N/A	4.649e+02	24.190	ug/g
7 SAMPLE	S98T002935	0 W @IC4G-01 FORMATE2	SOLID	N/A	2.446e+02	75.000	ug/g
8 DUP	S98T002935	0 W @IC4G-01 F*4-01	SOLID	7.77e+02	1.04e+03	28.949	RPD
8 DUP	S98T002935	0 W @IC4G-01 ACETATE2	SOLID	4.65e+02	5.69e+02	20.116	RPD
8 DUP	S98T002935	0 W @IC4G-01 FORMATE2	SOLID	2.45e+02	3.03e+02	21.168	RPD
8 DUP	S98T002935	0 W @IC4G-01 GLYCOLTI	SOLID	?	<6.76e1		RPD
9 SPK	S98T002935	0 W @IC4G-01 F*4-01	SOLID	6.10e1	6.89e+01	112.951	% Recovery
9 SPK	S98T002935	0 W @IC4G-01 ACETATE2	SOLID	1.16e2	1.12e+02	96.552	% Recovery
9 SPK	S98T002935	0 W @IC4G-01 FORMATE2	SOLID	1.19e2	1.19e+02	100.000	% Recovery
9 SPK	S98T002935	0 W @IC4G-01 GLYCOLTI	SOLID	1.08e2	0.00e+00	0.000	% Recovery

Final page for worklist# 27935

Analyst Signature Date Analyst Signature Date

Janis M. Lutz 2/22/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27935

Analyst: ADD Instrument: IC 45S2 Book# 59N21-A LC8
Method: LA-533-105 Rev/Mod F-0 58N21-A CCV

Worklist Comment: BY-105 RERUN DF=6

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC	QC		
2	LCS-INST				@IC4G-QC	QC		
3	CCV				@IC4G-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002934	0	W	@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2								
6	DUP	S98T002934	0	W	@IC4G-01	SOLID		
7	SAMPLE	S98T002935	0	W	@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2								
8	DUP	S98T002935	0	W	@IC4G-01	SOLID		
9	SPK	S98T002935	0	W	@IC4G-01	SOLID		

Final page for worklist # 27935

John W. Purnell 02-18-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 2-19-99
John W. Purnell

Validated 2/22/99 JMB

27935FEB.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 59N21-A LCS Date: 02/18/1999 20:30:40
 Data File : C:\DX\DATA\99021811.D04
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-2
 Analyst : *Anthony Parekh* Column: AG14A-SC, AS14A-SC, SRS

OK 02-18-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	20420	871628	1	
2	2.20		0.000	10465	269168	3	
7	2.89		0.000	2571	23653	2	
10	3.23		0.000	1569	24635	2	
11	3.65		0.000	1930	16008	2	
13	4.00		0.000	570	17956	1	
18	4.85		0.000	3694	23538	2	
19	4.97		0.000	3620	26297	2	
20	5.16		0.000	3233	22758	2	
21	5.29		0.000	3405	42563	2	
25	6.08		0.000	4591	38549	2	
26	6.25		0.000	3799	44001	2	
27	6.47		0.000	3822	49991	2	
28	7.41	fluoride	63.041	1911968	28787515	2	0.00
29	8.13	glycolate	105.933	517455	9620784	2	0.00
30	8.64	acetate	116.092	540499	10617547	2	0.00
31	9.36	formate	123.586	976086	18997624	2	0.00
Totals			408.652	4009698	69494214		

File: 99021811.D04 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS ANALYST'S SIGNATURE
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1619 TO 1637

```

=====
Sample Name: 58N21-A CCV                      Date: 02/18/1999 20:52:51
Data File  : C:\DX\DATA\99021811.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

----- Calibration Volume Dilution Points Rate Start Stop Area Reject -----

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	11992	460235	2	
3	2.19		0.000	23559	408214	2	
4	2.28		0.000	23056	301591	2	
5	2.55		0.000	18161	120864	2	
6	2.91		0.000	17844	440039	2	
7	3.05		0.000	15735	146749	2	
8	3.21		0.000	12434	60850	2	
9	3.31		0.000	10611	124989	2	
13	4.13		0.000	3282	82071	2	
21	6.37		0.000	3177	18002	2	
22	6.45		0.000	3087	41622	2	
23	7.41	fluoride	57.914	1769973	26330586	2	0.00
24	8.13	glycolate	96.342	477160	8714897	2	0.00
25	8.61	acetate	121.865	547218	11153438	2	0.00
26	9.36	formate	121.809	991342	18710390	2	0.00
Totals			397.929	3928632	67114536		

File: 99021811.D05 Sample: 58N21-A CCV

```

=====
Sample Name: BLANK                      Date: 02/18/1999 19:28:57
Data File  : C:\DX\DATA\99021811.D01
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 1      Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 5700 5Hz 0.00 19.00 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	3.33		0.000	270800	17822240	1	
2	4.59		0.000	14873	271541	2	
3	5.52		0.000	77895	5256459	3	
4	6.59		0.000	2685	50215	4	
5	6.85		0.000	7849	130368	4	
6	7.20	fluoride	0.027	11479	198968	4	0.00
7	7.52		0.000	13723	130706	4	
8	8.29		0.000	26721	1739908	2	
9	9.25	formate	0.050	6184	128553	2	0.00
10	9.73		0.000	2916	31322	2	
Totals			0.077	435126	25760280		

File: 99021811.D01 Sample: BLANK

```

=====
Sample Name: BLANK Prep                      Date: 02/18/1999 21:13:38
Data File  : C:\DX\DATA\99021811.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.73		0.000	802	38004	1	
5	2.92		0.000	3112	63813	2	
6	3.17		0.000	4063	26950	2	
7	3.29		0.000	3947	34524	2	
8	3.51		0.000	3289	20750	2	
9	3.59		0.000	2965	38839	2	
19	5.60		0.000	3229	77934	2	
20	6.00		0.000	2341	20031	2	
21	6.31		0.000	1542	26243	2	
22	6.69		0.000	2929	27925	2	
23	6.99		0.000	3546	81281	2	
24	7.39	fluoride	0.024	1398	30341	2	0.00
25	7.87		0.000	3116	87705	2	
26	8.59	acetate	0.028	2013	26470	2	0.00
27	9.04		0.000	2523	25137	2	
28	9.25	formate	0.047	3553	77402	2	0.00
Totals			0.098	44368	703349		

File: 99021811.D06 Sample: BLANK Prep

```

=====
Sample Name: S98T002934 SAM                      Date: 02/18/1999 21:34:52
Data File  : C:\DX\DATA\99021811.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7                Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	4049	71410	2	
2	0.73		0.000	1800	17321	2	
4	1.70		0.000	124558	1602520	1	
5	3.03		0.000	9948	262534	2	
6	3.09		0.000	9473	44678	2	
7	3.31		0.000	6463	99674	2	
8	3.45		0.000	4008	25574	2	
11	3.85		0.000	1494	22313	2	
12	4.15		0.000	5204	37078	2	
13	4.24		0.000	5195	46040	2	
14	4.40		0.000	5099	25316	2	
15	4.56		0.000	6103	103678	2	
16	4.96		0.000	5564	61611	2	
17	5.07		0.000	6893	67607	2	
18	5.23		0.000	7477	28954	2	
19	5.32		0.000	9388	81601	2	
20	5.47		0.000	9533	332017	2	
21	7.25	fluoride	3.961	1832406	30535104	3	0.00
22	8.00	glycolate	0.240	17921	193577	4	0.00
23	8.45	acetate	2.151	142809	3162409	2	0.00
24	9.20	formate	1.301	137966	2775474	2	0.00
25	10.00		0.000	18004	282060	1	
Totals			7.653	2371355	39878552		

File: 99021811.D07 Sample: S98T002934 SAM

```

=====
Sample Name: S98T002934 DUP                      Date: 02/18/1999 21:55:21
Data File  : C:\DX\DATA\99021811.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8                Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	5987	200226	1	
3	1.70		0.000	126359	1591565	2	
4	2.92		0.000	15240	290413	2	
5	3.07		0.000	42700	1325125	2	
6	3.87		0.000	15571	208552	2	
7	4.15		0.000	11853	90009	2	
8	4.33		0.000	8987	114405	2	
9	4.53		0.000	5600	50602	2	
10	4.73		0.000	2613	21196	2	
19	6.64		0.000	1948	16890	2	
20	7.25	fluoride	3.607	1668896	27673006	2	0.00
21	8.45	acetate	1.944	126852	2836215	2	0.00
22	9.17	formate	1.153	120446	2384846	2	0.00
23	10.00		0.000	16806	251946	1	

Totals 6.704 2169856 37054996

File: 99021811.D08 Sample: S98T002934 DUP

```

=====
Sample Name: S98T002935 SAM                      Date: 02/18/1999 22:16:12
Data File  : C:\DX\DATA\99021811.D09
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 9              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	5048	88704	1	
2	1.07		0.000	1978	25865	2	
4	1.70		0.000	107107	1310550	1	
5	2.65		0.000	8448	52879	2	
6	2.93		0.000	10668	167783	2	
7	3.05		0.000	10039	79119	2	
8	3.23		0.000	11327	64429	2	
9	3.29		0.000	10662	74195	2	
10	3.44		0.000	8792	75675	2	
11	3.64		0.000	6978	52981	2	
12	3.72		0.000	6632	43420	2	
13	3.85		0.000	3703	23821	2	
14	3.99		0.000	2793	18969	3	
17	4.47		0.000	3222	44231	2	
18	4.63		0.000	3149	19505	2	
19	4.81		0.000	3471	41975	2	
23	5.51		0.000	3869	40801	2	
24	5.79		0.000	6049	133883	2	
25	7.25	fluoride	3.854	1797206	29668966	2	0.00
26	8.45	acetate	2.306	148768	3407132	2	0.00
27	9.20	formate	1.213	125798	2542080	2	0.00
28	10.00		0.000	15932	239738	1	
Totals			7.372	2301639	38216702		

File: 99021811.D09 Sample: S98T002935 SAM

HNF-1663 REV. 0

Sample Name: S98T002935 DUP Date: 02/18/1999 22:36:32
 Data File : C:\DX\DATA\99021811.D10
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 10 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	3451	65316	1	
3	1.70		0.000	94288	1178886	1	
4	2.96		0.000	14187	306201	2	
5	3.08		0.000	13546	87848	2	
6	3.23		0.000	12227	71737	2	
7	3.29		0.000	11982	47677	2	
8	3.44		0.000	11583	120937	2	
9	3.60		0.000	9395	67961	2	
10	3.73		0.000	9225	91061	2	
11	3.91		0.000	7728	76565	2	
12	4.20		0.000	8130	95832	2	
13	4.28		0.000	7996	67230	2	
14	4.48		0.000	6681	88734	2	
15	4.77		0.000	2383	20460	2	
16	4.95		0.000	2186	18786	2	
18	5.21		0.000	2577	16748	2	
19	5.33		0.000	3103	23794	2	
20	5.49		0.000	5131	172202	2	
21	7.25	fluoride	5.196	2522111	40558770	2	0.00
22	8.45	acetate	2.830	188481	4233745	2	0.00
23	9.20	formate	1.507	169893	3322132	2	0.00
24	10.00		0.000	19445	304016	1	
Totals			9.532	3125730	51036635		

File: 99021811.D10 Sample: S98T002935 DUP

Sample Name: S98T002935 SPK

Date: 02/18/1999 22:56:12

Data File : C:\DX\DATA\99021811.D11

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 11

Detector: CDM-2

Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	6	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	3440	68600	1	
2	1.70		0.000	106166	1279172	1	
3	2.85		0.000	12419	252322	2	
4	3.03		0.000	11270	91646	2	
5	3.19		0.000	10222	141204	2	
6	3.51		0.000	7425	123327	2	
7	3.77		0.000	5750	82180	2	
8	3.96		0.000	3921	44226	2	
19	5.55		0.000	3765	36440	2	
20	5.65		0.000	4116	37923	2	
21	5.87		0.000	4038	86780	2	
22	7.25	fluoride	7.299	3559683	57827224	2	0.00
23	7.97	glycolate	5.332	418962	8095577	2	0.00
24	8.45	acetate	7.886	591747	12163158	2	0.00
25	9.17	formate	7.184	932926	18568701	2	0.00
26	10.00		0.000	14151	205922	1	
Totals			27.701	5690002	99104402		

File: 99021811.D11 Sample: S98T002935 SPK

LABCORE Completed Worklist Report for Worklist# 27991

Analyst: vlm

Instrument: IC45S2

Book#: 35N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 for @IC4G-01, try 21df. jmf

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4 QC	QC	1	<1.10e-2		ug/mL
1 CCB	0	@IC4G-QC ACETATE2 QC	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2 QC	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1 QC	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4 QC	QC	6.05e1	6.22e+01	102.810 % Recovery	
2 LCS-INST	0	@IC4G-QC ACETATE2 QC	QC	1.19e2	9.22e+01	77.479 % Recovery	
2 LCS-INST	0	@IC4G-QC FORMATE2 QC	QC	1.28e2	1.17e+02	91.406 % Recovery	
2 LCS-INST	0	@IC4G-QC GLYCOLT1 QC	QC	1.16e2	1.08e+02	93.103 % Recovery	
3 CCV	0	@IC4G-QC F*4 QC	QC	6.21e1	5.94e+01	95.652 % Recovery	
3 CCV	0	@IC4G-QC ACETATE2 QC	QC	1.23e2	1.16e+02	94.309 % Recovery	
3 CCV	0	@IC4G-QC FORMATE2 QC	QC	1.30e2	1.24e+02	95.385 % Recovery	
3 CCV	0	@IC4G-QC GLYCOLT1 QC	QC	1.02e2	9.90e+01	97.059 % Recovery	
4 BLNK-PREP	0	@IC4G-01 F*4-01 SOLID	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2 SOLID	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2 SOLID	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLT1 SOLID	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002931 0 W	@IC4G-01 F*4-01 SOLID	SOLID	N/A	1.031e+03	129.400	ug/g
5 SAMPLE	S98T002931 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	N/A	1.791e+02	83.480	ug/g
5 SAMPLE	S98T002931 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	N/A	< 2.588e+02	258.700	ug/g
5 SAMPLE	S98T002931 0 W	@IC4G-01 GLYCOLT1 SOLID	SOLID	N/A	< 2.338e+02	233.700	ug/g
6 DUP	S98T002931 0 W	@IC4G-01 F*4-01 SOLID	SOLID	1.03e+03	1.17e+03	12.727	RPD
6 DUP	S98T002931 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	1.79e+02	1.86e+02	3.836	RPD
6 DUP	S98T002931 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	<2.59e2	<2.53e2		RPD
6 DUP	S98T002931 0 W	@IC4G-01 GLYCOLT1 SOLID	SOLID	<2.34e2	<2.29e2		RPD
7 SPK	S98T002931 0 W	@IC4G-01 F*4-01 SOLID	SOLID	6.05e1	6.88e+01	113.719 % Recovery	
7 SPK	S98T002931 0 W	@IC4G-01 ACETATE2 SOLID	SOLID	1.19e2	9.13e+01	76.723 % Recovery	
7 SPK	S98T002931 0 W	@IC4G-01 FORMATE2 SOLID	SOLID	1.28e2	1.20e+02	93.750 % Recovery	
7 SPK	S98T002931 0 W	@IC4G-01 GLYCOLT1 SOLID	SOLID	1.16e2	1.06e+02	91.379 % Recovery	

Final page for worklist# 27991

Analyst Signature

Date

Analyst Signature

Date

John M. Lutz 1/28/99
Reviewer Signature Date

01/26/99 08:09

A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27991

Analyst: JMM Instrument: IC ✓Book# CCV ^{EC 5} ^{35 N21-A}
^{34 N21-A}Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 for @IC4G-01, try 21df. jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC QC			
2 LCS-INST			@IC4G-QC QC			
3 CCV			@IC4G-QC QC			
4 BLNK-PREP			@IC4G-01 SOLID			
5 SAMPLE	S98T002931 0 W		@IC4G-01 SOLID		98000490 BY-105 (2)	
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
6 DUP	S98T002931 0 W		@IC4G-01 SOLID			
7 SPK	S98T002931 0 W		@IC4G-01 SOLID			

Final page for worklist # 27991

Valerie Massin 01-26-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

uploaded 1-26-99John Worell27991JAN.CSVre uploaded 1/28/99validated 1/28/99 jmf

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

=====

Sample Name: CCB	Date: 01/26/1999 11:04:02
Data File : C:\DX\DATA\99012601.D02	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 2	Detector: CDM-2
Analyst : <i>Valerie Z. Masarik</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

01-26-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	4881	5Hz	0.00	16.27	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.50		0.000	18441	590177	2	
16	4.71		0.000	3954	55870	2	
31	8.19		0.000	11171	485771	2	
32	9.36		0.000	5723	121077	2	
Totals			0.000	39290	1252895		

File: 99012601.D02 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1533 TO 1689.

Sample Name: LCS 35N21-A Date: 01/26/1999 11:23:40
 Data File : C:\DX\DATA\99012601.D03
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-2
 Analyst : *Valerie Mason* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 4946 5Hz 0.00 16.48 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.43		0.000	15538	300323	2	
7	2.95		0.000	5838	77919	2	
12	3.83		0.000	5024	79407	2	
28	7.25	fluoride	101.14 62.199	1791668	26708225	2	0.00
29	7.95	glycolate	93.36 108.324	523849	9647412	2	0.00
30	8.43	acetate	77.43 92.207	463152	9279715	2	0.00
31	9.17	formate	91.69 117.364	1029678	19718219	2	0.00
Totals			380.094	3834746	65811220		

File: 99012601.D03 Sample: LCS 35N21-A

Sample Name: CCV 34N21-A Date: 01/26/1999 11:41:22
 Data File : C:\DX\DATA\99012601.D04
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-2
 Analyst : *Valerie Moore* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 4966 5Hz 0.00 16.55 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	4812	91156	1	
3	2.24		0.000	4721	57384	1	
5	2.57		0.000	4826	57285	2	
31	7.23	fluoride	95.63	59.386	1709362	2	0.00
32	7.92	glycolate	97.84	98.982	467127	2	0.00
33	8.40	acetate	93.91	115.515	588095	2	0.00
34	9.15	formate	95.15	123.698	1068434	2	0.00
Totals			397.582	3847379	66845419		

File: 99012601.D04 Sample: CCV 34N21-A

=====

Sample Name: PREP BLANK	Date: 01/26/1999 12:04:32
Data File : C:\DX\DATA\99012601.D05	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 5	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	50000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	4756	90844		1
Totals			0.000	4756	90844		

File: 99012601.D05 Sample: PREP BLANK

Data Reprocessed On 01/26/1999 13:15:35

```

=====
Sample Name: S98T002931                      Date: 01/26/1999 13:11:43
Data File  : C:\DX\DATA\99012601.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    4820  5Hz   0.00 16.06    50000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
6	2.23		0.000	5917	101753	2	
7	2.60		0.000	8079	56047	2	
8	2.72		0.000	9318	254669	2	
9	3.19		0.000	9431	122929	2	
10	3.45		0.000	8538	118779	2	
11	3.71		0.000	9581	339525	2	
14	4.87		0.000	6341	125032	2	
15	5.28		0.000	5916	73453	2	
17	6.03		0.000	3900	50569	2	
20	7.17	fluoride	5.186	671721	10396181	2	0.00
21	8.35	acetate	0.901	30933	649805	2	0.00
22	9.07	formate	0.941	31611	703709	3	0.00
24	9.87		0.000	8228	161287	2	
Totals			7.029	809515	13153737		

```

=====
Sample Name: S98T002931 DUP                      Date: 01/26/1999 13:31:15
Data File  : C:\DX\DATA\99012601.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7               Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 5173 5Hz 0.00 17.24 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	11534	110884	1	
4	2.77		0.000	6555	89923	2	
5	2.88		0.000	5483	64955	2	
6	3.11		0.000	5325	56148	2	
24	7.17	fluoride	6.023	766887	12142336	2	0.00
25	8.37	acetate	0.954	30236	675126	2	0.00
26	9.12	formate	0.815	28639	602600	2	0.00
27	9.89		0.000	6953	121496	1	
Totals			7.792	861612	13863467		

File: 99012601.D07 Sample: S98T002931 DUP

```

=====
Sample Name: S98T002931 SPIKE      Date: 01/26/1999 13:58:33
Data File  : C:\DX\DATA\99012601.D08
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 8 Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 4911 5Hz 0.00 16.37 50000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	2667	69742	1	
5	2.24		0.000	5989	100022	1	
14	3.83		0.000	5152	50175	3	
20	4.88		0.000	2911	51975	2	
25	5.91		0.000	5155	79042	2	
27	7.17	fluoride	18.947	2691628	39603486	2	0.00
28	7.87	glycolate	21.141	512216	9052081	2	0.00
29	8.32	acetate	19.155	482820	9271621	2	0.00
30	9.04	formate	24.062	1027173	19438557	3	0.00
Totals			83.304	4735711	77716701		

File: 99012601.D08 Sample: S98T002931 SPIKE

LABCORE Completed Worklist Report for Worklist# 28021

Analyst: adp

Instrument: IC40S1

Book#: 50N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 RERUN/HIGH HNO3,HIGH RPD/DF=41 JHW

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	0	0IC-QC	F	QC	1	<1.20e-2	ug/mL
1 CCB	0	0	0IC-QC	CL	QC	1	5.90e-02	0.059 ug/mL
1 CCB	0	0	0IC-QC	NO2	QC	1	<1.08e-1	ug/mL
1 CCB	0	0	0IC-QC	BR	QC	1	<1.25e-1	ug/mL
1 CCB	0	0	0IC-QC	NO3	QC	1	<1.39e-1	ug/mL
1 CCB	0	0	0IC-QC	P04	QC	1	<1.20e-1	ug/mL
1 CCB	0	0	0IC-QC	S04	QC	1	<1.38e-1	ug/mL
1 CCB	0	0	0IC-QC	OXALATE2	QC	1	<1.05e-1	ug/mL
2 LCS-INST	0	0	0IC-QC	F	QC	5.92e1	6.40e+01	108.108 % Recovery
2 LCS-INST	0	0	0IC-QC	CL	QC	7.96e1	8.42e+01	105.779 % Recovery
2 LCS-INST	0	0	0IC-QC	NO2	QC	5.35e2	4.85e+02	90.654 % Recovery
2 LCS-INST	0	0	0IC-QC	BR	QC	5.86e2	5.81e+02	99.147 % Recovery
2 LCS-INST	0	0	0IC-QC	NO3	QC	5.92e2	6.51e+02	109.966 % Recovery
2 LCS-INST	0	0	0IC-QC	P04	QC	5.48e2	5.71e+02	104.197 % Recovery
2 LCS-INST	0	0	0IC-QC	S04	QC	6.38e2	6.43e+02	100.784 % Recovery
2 LCS-INST	0	0	0IC-QC	OXALATE2	QC	5.52e2	6.04e+02	109.420 % Recovery
3 CCV	0	0	0IC-QC	F	QC	6.49e1	6.24e+01	97.500 % Recovery
3 CCV	0	0	0IC-QC	CL	QC	8.95e1	9.14e+01	102.123 % Recovery
3 CCV	0	0	0IC-QC	NO2	QC	5.37e2	5.24e+02	97.579 % Recovery
3 CCV	0	0	0IC-QC	BR	QC	6.30e2	6.05e+02	96.032 % Recovery
3 CCV	0	0	0IC-QC	NO3	QC	7.09e2	6.75e+02	96.429 % Recovery
3 CCV	0	0	0IC-QC	P04	QC	6.35e2	6.44e+02	101.417 % Recovery
3 CCV	0	0	0IC-QC	S04	QC	7.09e2	6.74e+02	96.286 % Recovery
3 CCV	0	0	0IC-QC	OXALATE2	QC	5.33e2	5.63e+02	105.629 % Recovery
4 BLNK-PREP	0	0	0IC-01	F-02	SOLID	1	<1.20e-2	ug/g
4 BLNK-PREP	0	0	0IC-01	CL-02	SOLID	1	<1.70e-2	ug/g
4 BLNK-PREP	0	0	0IC-01	NO2-02	SOLID	1	<1.08e-1	ug/g
4 BLNK-PREP	0	0	0IC-01	BR-02	SOLID	1	<1.25e-1	ug/g
4 BLNK-PREP	0	0	0IC-01	NO3-02	SOLID	1	<1.39e-1	ug/g
4 BLNK-PREP	0	0	0IC-01	P04-02	SOLID	1	<1.20e-1	ug/g
4 BLNK-PREP	0	0	0IC-01	S04-02	SOLID	1	<1.38e-1	ug/g
4 BLNK-PREP	0	0	0IC-01	OXALATE2	SOLID	1	<1.05e-1	ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	F-02	SOLID	N/A	2.981e+02	99.510 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	CL-02	SOLID	N/A	2.576e+02	141.000 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	NO2-02	SOLID	N/A	3.028e+03	895.400 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	BR-02	SOLID	N/A	1.036e+03	1036.000 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	P04-02	SOLID	N/A	2.528e+03	995.100 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	S04-02	SOLID	N/A	1.800e+03	1144.000 ug/g
5 SAMPLE	S98T002613	0 W	0IC-01	OXALATE2	SOLID	N/A	2.941e+03	870.400 ug/g
6 DUP	S98T002613	0 W	0IC-01	F-02	SOLID	2.98e+02	5.58e+02	60.748 RPD
6 DUP	S98T002613	0 W	0IC-01	CL-02	SOLID	2.58e+02	2.40e+02	7.229 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28021

Seq Type	Sample#	K A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002613	0 W	@IC-01	NO2-02	SOLID	3.03e+03	2.95e+03	2.676 RPD
6 DUP	S98T002613	0 W	@IC-01	BR-02	SOLID	<1.04e3	<1.02e3	RPD
6 DUP	S98T002613	0 W	@IC-01	NO3-02	SOLID	5.94e+05	5.73e+05	3.599 RPD
6 DUP	S98T002613	0 W	@IC-01	PO4-02	SOLID	2.53e+03	2.57e+03	1.569 RPD
6 DUP	S98T002613	0 W	@IC-01	SO4-02	SOLID	1.80e+03	1.50e+03	18.182 RPD
6 DUP	S98T002613	0 W	@IC-01	OXALATE2	SOLID	2.94e+03	2.62e+03	11.511 RPD
7 SAMPLE	S98T002615	0 W	@IC-01	SO4-02	SOLID	N/A	2.221e+03	1133.000 ug/g
7 SAMPLE	S98T002615	0 W	@IC-01	OXALATE2	SOLID	N/A	1.474e+04	861.800 ug/g
8 DUP	S98T002615	0 W	@IC-01	SO4-02	SOLID	2.22e+03	3.21e+03	36.464 RPD
8 DUP	S98T002615	0 W	@IC-01	OXALATE2	SOLID	1.47e+04	1.38e+04	6.316 RPD

Final page for worklist# 28021

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jean M. Lyle
Reviewer Signature Date 2/9/99

01/26/99 09:41
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 28021

Analyst: ADD Instrument: IC 4051 Book# 51N21-C LCS
SON21-C CCV
 Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 RERUN/HIGH HNO3, HIGH RPD/DF=41 JHW

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002613 O W	@IC-01	SOLID		98000436 BY-105 (2)	
	Analytes Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
6 DUP	S98T002613 O W	@IC-01	SOLID			
7 SAMPLE	S98T002615 O W	@IC-01	SOLID		98000436 BY-105 (2)	
	Analytes Requested: OXALATE2, SO4-02					
8 DUP	S98T002615 O W	@IC-01	SOLID			

Final page for worklist # 28021

Anthony Perunth 2-8-99
 Analyst Signature Date

_____
 Analyst Signature Date

Data Entry Comments:

Please mix - uploaded on 2/8/99
NO3 rejected in S98T002613 because it
was over cal curve. Rest validated 2/8/99 m.f.
Reanalyzed on Wt# 28318. (Bar 3/24/99)

1642

Sample Name: 51N21-C LCS Date: 02/08/1999 00:16:15
 Data File : C:\DX\DATA\99020701.D16
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 16 Detector: CDM-1
 Analyst : *Anthony Parviz* Column: AG4A/AS4A anion column

02-08-99
 Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	94	371	2	
2	0.98	fluoride	64.035/08.17	1970	9729	2	-1.01
3	1.43	chloride	84.215/05.80	2379	10375	1	-0.23
4	1.68	nitrite	485.006/06.66	8292	41062	1	-1.75
5	2.44	bromide	580.719/09.10	6156	33537	1	-4.69
6	2.75	nitrate	650.502/4.68	8048	50963	1	-0.48
7	4.02	phosphate	570.641/04.13	1551	18872	1	1.52
8	5.17	sulfate	642.512/00.71	4619	62202	1	3.26
9	6.72	oxalate	604.422/10.50	2552	41223	1	4.13
Totals			3682.051	35661	268334		

File: 99020701.D16 Sample: 51N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1143 TO 1160.

```

=====
Sample Name: 50N21-C   CCV                      Date: 02/08/1999 00:38:35
Data File  : C:\DX\DATA\99020701.D18
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1   System: 1   Inject#: 18          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	91	357	2	
2	0.99	fluoride	62.372	1864	9472	2	-0.34
3	1.43	chloride	91.394	2592	11272	1	-0.23
4	1.68	nitrite	524.339	8921	44402	1	-1.75
5	2.44	bromide	605.120	6521	34970	1	-4.69
6	2.75	nitrate	675.442	8348	52948	1	-0.48
7	4.03	phosphate	643.564	1777	21341	1	1.68
8	5.17	sulfate	673.951	4369	65295	1	3.26
9	6.72	oxalate	563.308	2259	38398	1	4.13
Totals			3839.490	36742	278456		

File: 99020701.D18 Sample: 50N21-C CCV

```

=====
Sample Name: BLANK                      Date: 02/08/1999 00:27:06
Data File  : C:\DX\DATA\99020701.D17
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 17    Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	37	152	1	
2	1.35	chloride	0.059	85	730	1	-5.36
3	2.79	nitrate	0.014	23	112	1	0.97
Totals			0.074	145	994		

```

=====
Sample Name: PREP BLANK                      Date: 02/08/1999 00:49:25
Data File  : C:\DX\DATA\99020701.D19
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 19         Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.27		0.000	57	529	1	
2	1.44	chloride	0.016	48	194	1	0.70
3	1.69	nitrite	0.041	87	353	1	-0.97
4	2.47	bromide	0.015	20	88	1	-3.65
5	2.80	nitrate	0.025	38	194	1	1.45
6	5.33	sulfate	0.043	26	412	1	6.45
Totals			0.140	276	1770		

File: 99020701.D19 Sample: PREP BLANK

Data Reprocessed On 02/08/1999 09:44:22

```

=====
Sample Name: S98T002613 SAM                      Date: 02/08/1999 01:00:24
Data File  : F:\DATA\99020701.D20
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 20              Detector:CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000  5Hz  0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	169	692	2	
2	0.99	fluoride	1.474	151	544	2	-0.34
3	1.43	chloride	1.274	89	382	1	-0.23
4	1.68	nitrite	14.974	616	3115	1	-1.75
5	2.45	bromide	0.755	24	106	1	-0.14
6	2.63	nitrate	2937.625	80356	704418	1	-0.13
7	4.05	phosphate	12.501	95	999	1	2.36
8	5.23	sulfate	8.900	165	2092	1	4.32
9	6.77	oxalate	14.541	141	2424	1	4.96
Totals			2992.044	81806	714771		

File: 99020701.D20 Sample: S98T002613 SAM

Data Reprocessed On 02/08/1999 09:45:37

Sample Name: S98T002613	DUP	Date: 02/08/1999 01:11:12
Data File : F:\DATA\99020701.D21		
Method : C:\DX\METHOD\KIT.MET		
ACI Address: 1	System: 1	Inject#: 21
Analyst :	Column: AG4A/AS4A anion column	Detector: CDM-1

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	41	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	191	904	2	
2	0.99	fluoride	2.814	201	1038	2	-0.34
3	1.43	chloride	1.212	84	363	1	-0.23
4	1.68	nitrite	14.888	619	3097	1	-1.75
5	2.63	nitrate	2889.730	78483	688988	1	-0.13
6	4.05	phosphate	12.956	98	1036	1	2.36
7	5.23	sulfate	7.545	135	1773	1	4.32
8	6.77	oxalate	13.203	122	2201	1	4.96
Totals			2942.347	79933	699401		

File: 99020701.D21 Sample: S98T002613 DUP

```

=====
Sample Name: S98T002615 SAM                      Date: 02/08/1999 01:21:53
Data File  : F:\DATA\99020701.D22
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 22              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          41    3000  5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.83		0.000	106	541	2	
2	0.99	fluoride	1.558	67	575	2	0.34
3	1.41	chloride	8.795	620	2642	1	-1.63
4	1.66	nitrite	148.469	6124	30943	1	-2.92
5	2.71	nitrate	272.804	8141	52677	1	0.12
6	4.00	phosphate	126.877	839	10241	1	1.01
7	5.17	sulfate	11.093	196	2607	1	3.26
8	6.72	oxalate	73.611	764	12296	1	4.13
Totals			643.206	16856	112522		

=====

Sample Name: S98T002615	DUP	Date: 02/08/1999 01:34:37
Data File : C:\DX\DATA\99020701.D23		
Method : C:\DX\METHOD\KIT.MET		
ACI Address: 1	System: 1	Inject#: 23
Analyst :	Column: AG4A/AS4A anion column	Detector: CDM-1

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	41	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	98	493	1	
2	1.02	fluoride	0.267	32	99	1	3.03
3	1.43	chloride	10.050	690	3020	1	-0.23
4	1.68	nitrite	168.471	6914	35122	1	-1.75
5	2.75	nitrate	274.244	8139	52959	1	-0.48
6	4.04	phosphate	90.638	603	7293	1	2.02
7	5.23	sulfate	16.520	302	3884	1	4.32
8	6.77	oxalate	71.132	766	11881	1	4.96
Totals			631.322	17545	114752		

File: 99020701.D23 Sample: S98T002615 DUP

LABCORE Completed Worklist Report for Worklist# 28066

Analyst: adp

Instrument: IC40S1

Book#: 51N21C

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	0	01C-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	0	01C-QC CL	QC	1	2.30e-02	0.023	ug/mL
1 CCB	0	0	01C-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	0	01C-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	0	01C-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	0	01C-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	0	01C-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	0	01C-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	0	01C-QC F	QC	5.92e1	6.49e+01	109.628 % Recovery	
2 LCS-INST	0	0	01C-QC CL	QC	7.96e1	7.72e+01	96.985 % Recovery	
2 LCS-INST	0	0	01C-QC NO2	QC	5.35e2	4.55e+02	85.047 % Recovery	
2 LCS-INST	0	0	01C-QC BR	QC	5.86e2	5.59e+02	95.392 % Recovery	
2 LCS-INST	0	0	01C-QC NO3	QC	5.92e2	6.32e+02	106.757 % Recovery	
2 LCS-INST	0	0	01C-QC P04	QC	5.48e2	5.42e+02	98.905 % Recovery	
2 LCS-INST	0	0	01C-QC S04	QC	6.38e2	6.21e+02	97.335 % Recovery	
2 LCS-INST	0	0	01C-QC OXALATE2	QC	5.52e2	5.86e+02	106.159 % Recovery	
3 CCV	0	0	01C-QC F	QC	6.40e1	6.29e+01	98.281 % Recovery	
3 CCV	0	0	01C-QC CL	QC	8.95e1	8.60e+01	96.089 % Recovery	
3 CCV	0	0	01C-QC NO2	QC	5.37e2	5.07e+02	94.413 % Recovery	
3 CCV	0	0	01C-QC BR	QC	6.30e2	6.07e+02	96.349 % Recovery	
3 CCV	0	0	01C-QC NO3	QC	7.00e2	6.74e+02	96.286 % Recovery	
3 CCV	0	0	01C-QC P04	QC	6.35e2	6.34e+02	99.843 % Recovery	
3 CCV	0	0	01C-QC S04	QC	7.00e2	6.52e+02	93.143 % Recovery	
3 CCV	0	0	01C-QC OXALATE2	QC	5.33e2	5.46e+02	102.439 % Recovery	
4 BLNK-PREP	0	0	01C-01 F-02	SOLID	1	<1.20e-2		ug/g
4 BLNK-PREP	0	0	01C-01 CL-02	SOLID	1	<1.70e-2		ug/g
4 BLNK-PREP	0	0	01C-01 NO2-02	SOLID	1	<1.08e-1		ug/g
4 BLNK-PREP	0	0	01C-01 BR-02	SOLID	1	<1.25e-1		ug/g
4 BLNK-PREP	0	0	01C-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4 BLNK-PREP	0	0	01C-01 P04-02	SOLID	1	<1.20e-1		ug/g
4 BLNK-PREP	0	0	01C-01 S04-02	SOLID	1	<1.38e-1		ug/g
4 BLNK-PREP	0	0	01C-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5 SAMPLE	S98T003390	0 W	01C-01 P04-02	SOLID	N/A	4.159e+03	2407.000	ug/g
6 DUP	S98T003390	0 W	01C-01 P04-02	SOLID	4.16e+03	3.88e+03	6.965	RPD
7 SPK	S98T003390	0 W	01C-01 P04-02	SOLID	5.48e2	5.58e+02	101.825 % Recovery	
8 SAMPLE	S98T003400	0 W	01C-01 NO3-02	SOLID	N/A	6.384e+04	576.000	ug/g
8 SAMPLE	S98T003400	0 W	01C-01 S04-02	SOLID	N/A	4.134e+03	571.800	ug/g
8 SAMPLE	S98T003400	0 W	01C-01 OXALATE2	SOLID	N/A	2.045e+04	435.100	ug/g
9 DUP	S98T003400	0 W	01C-01 NO3-02	SOLID	6.38e+04	5.29e+04	18.680	RPD
9 DUP	S98T003400	0 W	01C-01 S04-02	SOLID	4.13e+03	2.57e+04	144.620	RPD
9 DUP	S98T003400	0 W	01C-01 OXALATE2	SOLID	2.04e+04	5.23e+04	87.758	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28066

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
10	SPK	S98T003400	0	W	@IC-01	NO3-02	SOLID	5.92e2	7.15e+02	120.777 % Recovery
10	SPK	S98T003400	0	W	@IC-01	SO4-02	SOLID	6.38e2	6.30e+02	98.746 % Recovery
10	SPK	S98T003400	0	W	@IC-01	OXALATE2	SOLID	5.52e2	6.11e+02	110.688 % Recovery

Final page for worklist# 28066

_____
Analyst Signature   Date

_____
Analyst Signature   Date

_____
Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 28066

Analyst: ADP Instrument: IC 4051 Book# 51N21-C LCS
Method: LA-533-105 Rev/Mod F-0 SON21-C CCV
Worklist Comment: BY105 FOR @IC-01 RLC

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T003390	0	W	@IC-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: PO4-02						
6	DUP	S98T003390	0	W	@IC-01	SOLID		
7	SPK	S98T003390	0	W	@IC-01	SOLID		
8	SAMPLE	S98T003400	0	W	@IC-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: NO3-02 , OXALATE2, SO4-02						
9	DUP	S98T003400	0	W	@IC-01	SOLID		
10	SPK	S98T003400	0	W	@IC-01	SOLID		

Final page for worklist # 28066

Anthony Lunn 2-8-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

Reverse Mix - data entered on 2/8/99
Validated 2/8/99 JMI Lye

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 51N21-C LCS Date: 02/08/1999 01:56:48
 Data File : C:\DX\DATA\99020701.D25
 Method : C:\DX\METHOD\KIT.MRT
 ACI Address: 1 System: 1 Inject#: 25 Detector: CDM-1
 Analyst : *Antony Purnell* Column: AG4A/AS4A anion column

02-08-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	76	285	2	
2	0.99	fluoride	64.903109.63	1910	9863	2	-0.34
3	1.43	chloride	77.18496.97	2249	9497	1	0.23
4	1.69	nitrite	455.41485.12	7875	38550	1	-1.36
5	2.45	bromide	558.53595.31	5921	32235	1	-4.43
6	2.75	nitrate	632.274106.80	7747	49514	1	-0.24
7	4.04	phosphate	541.82598.87	1504	17900	1	2.02
8	5.23	sulfate	620.74097.29	5164	60063	1	4.32
9	6.72	oxalate	585.736106.11	2368	39939	1	4.13
Totals			3536.610	34814	257846		

File: 99020701.D25 Sample: 51N21-C LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1654 TO 1663.

Sample Name: 50N21-C CCV Date: 02/08/1999 02:07:57
 Data File : C:\DX\DATA\99020701.D26
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 26 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	79	301	2	
2	1.01	fluoride	62.910	1866	9555	2	1.68
3	1.45	chloride	85.997	2459	10597	1	1.17
4	1.71	nitrite	506.588	8695	42894	1	-0.19
5	2.47	bromide	606.969	6297	35078	1	-3.39
6	2.78	nitrate	673.823	8156	52820	1	0.72
7	4.07	phosphate	634.346	1736	21028	1	2.69
8	5.23	sulfate	651.842	4788	63120	1	4.32
9	6.77	oxalate	545.736	2309	37191	1	4.96
Totals			3768.211	36386	272584		

File: 99020701.D26 Sample: 50N21-C CCV

```

=====
Sample Name: BLANK                      Date: 02/08/1999 01:45:15
Data File  : C:\DX\DATA\99020701.D24
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 24    Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.83		0.000	14	53	1	
2	1.42	chloride	0.023	33	277	1	-0.70
Totals			0.023	46	330		

```

=====
Sample Name: PREP BLANK                      Date: 02/08/1999 02:19:11
Data File  : C:\DX\DATA\99020701.D27
Method     : C:\DX\METHOD\RIT.MET
ACI Address: 1 System: 1 Inject#: 27         Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.27		0.000	57	512	1	
2	1.44	chloride	0.015	41	188	1	0.70
3	1.70	nitrite	0.009	14	73	1	-0.58
4	2.79	nitrate	0.008	14	63	1	0.97
Totals			0.032	126	836		

File: 99020701.D27 Sample: PREP BLANK

Sample Name: S98T003390 SAM	Date: 02/08/1999 02:41:56
Data File : F:\DATA\99020701.D29	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 29	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	76	303	2	
2	0.99	fluoride	5.467	193	819	2	-0.34
3	1.43	chloride	1.974	59	240	1	-0.23
4	1.68	nitrite	27.042	470	2283	1	-1.75
5	2.69	nitrate	3226.326	36378	270500	1	-0.12
6	4.05	phosphate	20.947	65	679	1	2.19
7	5.23	sulfate	27.041	207	2580	1	4.32
8	6.77	oxalate	17.014	71	1151	1	4.96
Totals			3325.812	37520	278556		

File: 99020701.D29 Sample: S98T003390 SAM

Data Reprocessed On 02/08/1999 08:56:40

```

=====
Sample Name: S98T003390  DUP                      Date: 02/08/1999 02:52:37
Data File  : F:\DATA\99020701.D30
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 30              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	76	299	2	
2	0.99	fluoride	6.022	207	902	2	-0.34
3	1.43	chloride	1.923	58	234	1	0.23
4	1.68	nitrite	26.605	465	2247	1	-1.75
5	2.69	nitrate	3211.671	36505	269156	1	-0.12
6	4.04	phosphate	19.893	64	645	1	2.02
7	5.23	sulfate	26.941	212	2570	1	4.32
8	6.77	oxalate	15.815	69	1070	1	4.96
Totals			3308.870	37657	277123		

File: 99020701.D30 Sample: S98T003390 DUP

Data Reprocessed On 02/08/1999 08:58:08

```

=====
Sample Name: S98T003390  SPK                      Date: 02/08/1999 03:03:48
Data File  : F:\DATA\99020701.D31
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1  System: 1  Inject#: 31             Detector:CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101      3000  5Hz  0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	140	525	2	
2	0.99	fluoride	66.775	2109	10152	2	-0.34
3	1.42	chloride	78.643	2183	9679	1	-0.70
4	1.68	nitrite	480.434	8184	40674	1	-1.75
5	2.43	bromide	549.757	6069	31721	1	0.14
6	2.67	nitrate	3792.096	42764	323403	1	0.12
7	4.03	phosphate	578.565	1569	19140	1	1.68
8	5.17	sulfate	640.382	4451	61993	1	3.26
9	6.72	oxalate	593.269	2531	40457	1	4.13
Totals			6779.922	70000	537743		

File: 99020701.D31 Sample: S98T003390 SPK

```

=====
Sample Name: S98T003400 SAM                      Date: 02/08/1999 03:42:17
Data File  : F:\DATA\99020701.D34
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 34              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    3000   5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. %Delta Code
1	0.85		0.000	192	899	1
2	1.01	fluoride	0.292	74	210	1 1.68
3	1.42	chloride	9.170	1206	5398	1 -0.70
4	1.68	nitrite	167.697	13108	68410	1 -1.75
5	2.72	nitrate	323.546	18484	124655	1 0.00
6	4.03	phosphate	68.718	901	10836	1 1.68
7	5.23	sulfate	20.950	798	9630	1 4.32
8	6.72	oxalate	103.616	2102	33939	1 4.13
Totals			693.988	36866	253977	

File: 99020701.D34 Sample: S98T003400 SAM

Data Reprocessed On 02/08/1999 09:02:19

```

=====
Sample Name: S98T003400   DUP                               Date: 02/08/1999 03:53:00
Data File  : F:\DATA\99020701.D35
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1   System: 1   Inject#: 35                   Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    3000   5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	239	1113	1	
2	1.01	fluoride	0.198	50	143	1	1.68
3	1.41	chloride	7.611	987	4475	1	-1.17
4	1.67	nitrite	137.929	11048	56220	1	-2.14
5	2.72	nitrate	268.471	15304	102749	1	0.00
6	4.02	phosphate	61.446	808	9678	1	1.52
7	5.17	sulfate	130.262	4857	60629	1	3.26
8	6.67	oxalate	265.205	5431	87813	1	3.31
Totals			871.123	38724	322818		

File: 99020701.D35 Sample: S98T003400 DUP

```

=====
Sample Name: S98T003400 SPK                      Date: 02/08/1999 04:03:55
Data File  : C:\DX\DATA\99020701.D36
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 36              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    3000  SHz    0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	234	1001	2	
2	0.99	fluoride	11.895	1705	8677	2	0.34
3	1.43	chloride	25.488	3214	15199	1	-0.23
4	1.69	nitrite	268.446	21097	109815	1	-1.36
5	2.45	bromide	109.444	5752	30352	1	-4.43
6	2.71	nitrate	466.548	26305	183013	1	-1.69
7	4.03	phosphate	182.452	2439	29353	1	1.68
8	5.17	sulfate	146.888	4695	68499	1	3.26
9	6.72	oxalate	225.716	4770	74538	1	4.13
Totals			1436.876	70210	520445		

File: 99020701.D36 Sample: S98T003400 SPK

LABCORE Completed Worklist Report for Worklist# 28251

Analyst: adp

Instrument: IC45S2

Book#: 58N21A

Method: LA-533-105 Rev/Mod F-O

Worklist Comment: Was 28022 @ic4g-01 BY-105

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	9.20e-02	0.092	ug/mL
1 CCB	0		@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0		@IC4G-QC F*4	QC	6.10e1	6.61e+01	100.361 % Recovery	
2 LCS-INST	0		@IC4G-QC ACETATE2	QC	1.16e2	1.12e+02	96.552 % Recovery	
2 LCS-INST	0		@IC4G-QC FORMATE2	QC	1.19e2	1.17e+02	98.319 % Recovery	
2 LCS-INST	0		@IC4G-QC GLYCOLT1	QC	1.08e2	1.06e+02	98.148 % Recovery	
3 CCV	0		@IC4G-QC F*4	QC	6.29e1	6.34e+01	100.795 % Recovery	
3 CCV	0		@IC4G-QC ACETATE2	QC	1.20e2	1.20e+02	100.000 % Recovery	
3 CCV	0		@IC4G-QC FORMATE2	QC	1.20e2	1.21e+02	100.833 % Recovery	
3 CCV	0		@IC4G-QC GLYCOLT1	QC	1.03e2	1.04e+02	100.971 % Recovery	
4 BLNK-PREP	0		@IC4G-01 F*4-01	SOLID	1	3.20e-02	0.032	ug/g
4 BLNK-PREP	0		@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 FORMATE2	SOLID	1	6.20e-02	0.062	ug/g
4 BLNK-PREP	0		@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002612	0 W	@IC4G-01 ACETATE2	SOLID	N/A	2.301e+02	22.900	ug/g
5 SAMPLE	S98T002612	0 W	@IC4G-01 FORMATE2	SOLID	N/A	2.106e+02	70.980	ug/g
6 DUP	S98T002612	0 W	@IC4G-01 F*4-01	SOLID	?	7.85e+03		RPD
6 DUP	S98T002612	0 W	@IC4G-01 ACETATE2	SOLID	2.30e+02	3.09e+02	29.314	RPD
6 DUP	S98T002612	0 W	@IC4G-01 FORMATE2	SOLID	2.11e+02	2.63e+02	21.941	RPD
6 DUP	S98T002612	0 W	@IC4G-01 GLYCOLT1	SOLID	?	<6.43e1		RPD

Final page for worklist# 28251

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

John M. Fuge 2/17/99

99020511.D12
P13
D14
D15
D16
D17

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 28251

Analyst: APP Instrument: IC _____ Book# 58N21 A LCS
Method: LA-533-105 Rev/Mod FO 58N21 A CCV
Worklist Comment: Was 28022 @ic4g-01 BY-105

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002612	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, FORMATE2						
6 DUP	S98T002612	0 W	@IC4G-01	SOLID		

Final page for worklist # 28251

See Attached
Signature Date

Signature Date

28022 Feb.CSV

Data Entry Comments: RPD still high on acetate & formate in
this run validated 2/9/99 JM Lyle

LABCORE Data Entry Template for Worklist# 28022

Analyst: AOD Instrument: IC _____ Book# 59N21-A LCS
Method: LA-533-105 Rev/Mod F-0 58N21-A CCV

Worklist Comment: BY105/HIGH RPD/ DF=6 JHW

S	Type	Sample#	R	A	Test	Matrix	Group#	Project

1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002612	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, FORMATE2								
6	DUP	S98T002612	0	W	@IC4G-01	SOLID		

WRONG Tray → Become WKL 28251 to correct Tray error

Final page for worklist # 28022

Anthony Perre 2-6-99
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1666

Sample Name: 59N21-A LCS Date: 02/06/1999 03:06:57
 Data File : C:\DX\DATA\99020511.D13
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 13 Detector: CDM-2
 Analyst: *Anthony Purnath* Column: AG14A-SC, AS14A-SC, SRS

2-6-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.17		0.000	4997	181572	1	
2	1.27		0.000	2264	31618	1	
3	2.27		0.000	1561	17492	2	
5	2.83		0.000	2880	32827	2	
6	3.04		0.000	6462	50009	2	
7	3.13		0.000	6013	57222	2	
8	3.31		0.000	6318	49339	2	
9	3.45		0.000	5986	40286	2	
10	3.63		0.000	4450	31459	2	
11	3.77		0.000	4618	43109	2	
12	3.93		0.000	5704	20035	2	
13	4.04		0.000	5714	128686	2	
14	4.65		0.000	3481	45581	2	
15	4.84		0.000	4405	18698	2	
16	5.03		0.000	4988	77614	2	
17	5.24		0.000	4490	25820	2	
18	5.41		0.000	2254	16657	2	
19	5.60		0.000	2518	19385	2	
20	5.81		0.000	6400	114050	2	
23	6.67		0.000	4043	20665	2	
24	7.25	fluoride	66.083148,33	1962712	28641784	2	0.00
25	7.95	glycolate	106.35898.48	522089	9536739	2	0.00
26	8.45	acetate	112.04496.57	544815	10532586	2	0.00
27	9.15	formate	116.61897.99	1004192	19087411	2	0.00
Totals			401.104	4123353	68820646		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1667 TO 1674.

1668

Sample Name: 58N21-A CCV Date: 02/06/1999 03:27:19
 Data File : C:\DX\DATA\99020511.D14
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 14 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	2361	30500	1	
4	1.47		0.000	2960	30032	1	
7	2.51		0.000	2261	37778	2	
9	3.31		0.000	3623	52959	2	
10	3.52		0.000	2851	32387	2	
12	3.96		0.000	1853	17057	2	
13	4.27		0.000	4639	99022	2	
14	4.88		0.000	2043	28892	2	
15	5.09		0.000	4728	53651	2	
16	5.39		0.000	3353	43054	2	
20	5.95		0.000	4068	54804	2	
21	6.27		0.000	3426	36632	2	
22	7.20	fluoride	63.364	1817494	27396521	2	0.00
23	7.92	glycolate	104.422	509053	9352842	2	0.00
24	8.40	acetate	119.867	577825	11282238	2	0.00
25	9.12	formate	121.378	1051384	19918543	2	0.00
Totals			409.031	3993922	68466911		

File: 99020511.D14 Sample: 58N21-A CCV

```

=====
Sample Name: BLANK                      Date: 02/06/1999 02:45:12
Data File  : C:\DX\DATA\99020511.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12    Detector: CDM-2
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	15464	176920	1	
2	2.03		0.000	17481	300175	2	
3	2.24		0.000	8294	46731	2	
5	3.13		0.000	2661	56780	2	
9	4.40		0.000	2150	21520	2	
10	5.07		0.000	21706	573380	1	
11	6.11		0.000	8414	161020	2	
20	8.61		0.000	19281	431169	2	
21	9.17	formate	0.092	21593	545215	2	0.00
22	9.49		0.000	32570	449489	2	
23	9.73		0.000	36704	1264960	2	
Totals			0.092	186319	4027360		

File: 99020511.D12 Sample: BLANK

1670

```

=====
Sample Name: PREP BLANK          Date: 02/06/1999 03:47:31
Data File  : C:\DX\DATA\99020511.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 15      Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1          1    5700 5Hz    0.00 19.00      16000
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	4063	54148	1	
3	2.20		0.000	31570	575748	2	
4	2.29		0.000	31235	282451	2	
5	2.88		0.000	33754	939592	2	
6	3.00		0.000	33769	503000	2	
7	3.27		0.000	30489	199260	2	
8	3.35		0.000	30659	361262	2	
9	3.61		0.000	31019	223360	2	
10	3.69		0.000	31109	319213	2	
11	3.87		0.000	28924	298067	2	
12	4.08		0.000	25143	321224	2	
13	4.35		0.000	24123	178753	2	
14	4.45		0.000	27063	373612	2	
15	4.68		0.000	25698	395463	2	
16	4.97		0.000	21168	220974	2	
17	5.23		0.000	20713	216779	2	
18	5.35		0.000	22507	321939	2	
19	5.61		0.000	19947	271733	2	
20	5.97		0.000	15052	229645	2	
21	6.24		0.000	16452	252291	2	
22	6.48		0.000	14402	417109	2	
23	7.15	fluoride	0.032	11671	195806	2	0.00
24	7.39		0.000	9481	161211	2	
25	7.76		0.000	5301	73909	2	
26	8.03	glycolate	0.055	6121	142368	3	0.00
28	8.91		0.000	3131	32948	2	
29	9.20	formate	0.062	3169	43606	2	0.00
Totals			0.149	557731	7605474		

File: 99020511.D15 Sample: PREP BLANK

1672

```

=====
Sample Name: S98T002612 SAM                      Date: 02/06/1999 04:08:00
Data File  : C:\DX\DATA\99020511.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1           6    5700 5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	3176	37438	1	
2	1.70		0.000	131857	1685234	1	
3	2.95		0.000	15420	405830	2	
4	3.35		0.000	10831	241957	2	
5	3.51		0.000	11291	158212	2	
6	3.79		0.000	7777	89992	2	
7	4.40		0.000	2195	45937	2	
9	5.05		0.000	1732	17969	2	
10	5.31		0.000	4720	74423	2	
11	5.59		0.000	4266	45333	2	
12	5.84		0.000	2494	38427	2	
14	7.15	fluoride	36.062	17784151	311223397	3	0.00
15	8.29	acetate	1.206	74014	1654574	4	0.00
16	8.99	formate	1.104	101000	2143396	2	0.00
17	9.76		0.000	21352	453904	2	
Totals			38.372	18176276	318316023		

File: 99020511.D16 Sample: S98T002612 SAM

1673

```

=====
Sample Name: S98T002612  DUP                               Date: 02/06/1999 04:28:12
Data File  : C:\DX\DATA\99020511.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17                      Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6    5700  5Hz   0.00 19.00   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	5317	104196	1	
2	0.90		0.000	426	18200	1	
3	1.70		0.000	149223	1911372	1	
4	3.01		0.000	13233	466420	2	
5	3.43		0.000	4518	85334	2	
8	4.57		0.000	4360	60396	2	
9	4.88		0.000	3600	43420	2	
10	5.21		0.000	3160	40712	2	
11	5.45		0.000	5080	43140	2	
12	5.61		0.000	5020	107929	3	
14	7.15	fluoride	41.019	20649567	363485608	2	0.00
15	8.27	acetate	1.613	102659	2323054	2	0.00
16	8.99	formate	1.375	142897	2923487	2	0.00
17	9.76		0.000	24626	494552	2	
Totals			44.007	21113687	372107820		

File: 99020511.D17 Sample: S98T002612 DUP

1674

LABCORE Completed Worklist Report for Worklist# 28270

Analyst: adp

Instrument: IC40S1

Book#: 60N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 RERUN TRY D.F.41 HIGH RFD

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1 CCB	0	@IC-QC CL	QC	1	8.30e-02	0.083	ug/mL
1 CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1 CCB	0	@IC-QC S04	QC	1	<1.38e-1		ug/mL
1 CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	@IC-QC F	QC	5.920e1	6.45e+01	108.953 % Recovery	
2 LCS-INST	0	@IC-QC CL	QC	7.960e1	8.22e+01	103.266 % Recovery	
2 LCS-INST	0	@IC-QC NO2	QC	5.480e2	4.96e+02	90.511 % Recovery	
2 LCS-INST	0	@IC-QC BR	QC	5.860e2	5.70e+02	97.270 % Recovery	
2 LCS-INST	0	@IC-QC NO3	QC	5.920e2	5.93e+02	100.169 % Recovery	
2 LCS-INST	0	@IC-QC P04	QC	5.480e2	5.44e+02	99.270 % Recovery	
2 LCS-INST	0	@IC-QC S04	QC	6.380e2	6.27e+02	98.276 % Recovery	
2 LCS-INST	0	@IC-QC OXALATE2	QC	5.530e2	6.01e+02	108.680 % Recovery	
3 CCV	0	@IC-QC F	QC	6.40e1	6.21e+01	97.031 % Recovery	
3 CCV	0	@IC-QC CL	QC	8.95e1	8.95e+01	100.000 % Recovery	
3 CCV	0	@IC-QC NO2	QC	5.54e2	5.66e+02	102.166 % Recovery	
3 CCV	0	@IC-QC BR	QC	6.30e2	6.09e+02	96.667 % Recovery	
3 CCV	0	@IC-QC NO3	QC	7.00e2	6.70e+02	95.714 % Recovery	
3 CCV	0	@IC-QC P04	QC	6.35e2	6.30e+02	99.213 % Recovery	
3 CCV	0	@IC-QC S04	QC	7.00e2	6.73e+02	96.143 % Recovery	
3 CCV	0	@IC-QC OXALATE2	QC	5.49e2	5.87e+02	106.922 % Recovery	
4 BLNK-PREP	0	@IC-01 F-02	SOLID	1	<1.20e-2		ug/g
4 BLNK-PREP	0	@IC-01 CL-02	SOLID	1	<1.70e-2		ug/g
4 BLNK-PREP	0	@IC-01 NO2-02	SOLID	1	<1.08e-1		ug/g
4 BLNK-PREP	0	@IC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4 BLNK-PREP	0	@IC-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4 BLNK-PREP	0	@IC-01 P04-02	SOLID	1	<1.20e-1		ug/g
4 BLNK-PREP	0	@IC-01 S04-02	SOLID	1	<1.38e-1		ug/g
4 BLNK-PREP	0	@IC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5 SAMPLE	S98T002849	0 W @IC-01 F-02	SOLID	N/A	1.999e+03	243.300	ug/g
5 SAMPLE	S98T002849	0 W @IC-01 P04-02	SOLID	N/A	1.735e+04	2433.000	ug/g
6 DUP	S98T002849	0 W @IC-01 F-02	SOLID	2.00e+03	7.47e+02	91.227	RFD
6 DUP	S98T002849	0 W @IC-01 CL-02	SOLID	?	1.67e+03		RFD
6 DUP	S98T002849	0 W @IC-01 NO2-02	SOLID	?	1.81e+04		RFD
6 DUP	S98T002849	0 W @IC-01 BR-02	SOLID	?	<2.50e3		RFD
6 DUP	S98T002849	0 W @IC-01 NO3-02	SOLID	?	3.54e+05		RFD
6 DUP	S98T002849	0 W @IC-01 P04-02	SOLID	1.74e+04	5.22e+03	107.692	RFD
6 DUP	S98T002849	0 W @IC-01 S04-02	SOLID	?	<2.76e3		RFD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28270

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002849	0 W	@IC-01	OXALATE2	SOLID	?	4.42e+03	RPD
7 SAMPLE	S98T002931	0 W	@IC-01	CL-02	SOLID	N/A	7.227e+02	341.300 ug/g
7 SAMPLE	S98T002931	0 W	@IC-01	PO4-02	SOLID	N/A	4.465e+03	2409.000 ug/g
7 SAMPLE	S98T002931	0 W	@IC-01	SO4-02	SOLID	N/A	3.388e+03	2770.000 ug/g
7 SAMPLE	S98T002931	0 W	@IC-01	OXALATE2	SOLID	N/A	4.165e+03	2108.000 ug/g
8 DUP	S98T002931	0 W	@IC-01	F-02	SOLID	?	8.04e+02	RPD
8 DUP	S98T002931	0 W	@IC-01	CL-02	SOLID	7.23e+02	8.06e+02	10.857 RPD
8 DUP	S98T002931	0 W	@IC-01	NO2-02	SOLID	?	7.71e+03	RPD
8 DUP	S98T002931	0 W	@IC-01	BR-02	SOLID	?	<2.46e3	RPD
8 DUP	S98T002931	0 W	@IC-01	NO3-02	SOLID	?	5.86e+05	RPD
8 DUP	S98T002931	0 W	@IC-01	PO4-02	SOLID	4.46e+03	4.81e+03	7.551 RPD
8 DUP	S98T002931	0 W	@IC-01	SO4-02	SOLID	3.39e+03	3.01e+03	11.875 RPD
8 DUP	S98T002931	0 W	@IC-01	OXALATE2	SOLID	4.16e+03	3.48e+03	17.801 RPD

Final page for worklist# 28270

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28270

Analyst: ADP Instrument: IC 4051 Book# 61N21-B LCS
Method: LA-533-105 Rev/Mod F-0 60N21-A CCV

Worklist Comment: BY-105 RERUN TRY D.F.41 HIGH RFD

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002849 0 W		@IC-01	SOLID	98000437	BY-105 (2)
	Analytes Requested: F-02			, PO4-02		
6 DUP	S98T002849 0 W		@IC-01	SOLID		
7 SAMPLE	S98T002931 0 W		@IC-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: CL-02			, OXALATE2, PO4-02		, SO4-02
8 DUP	S98T002931 0 W		@IC-01	SOLID		

Final page for worklist # 28270

Anthony P. Smith 02-14-99
Signature Date

_____
Signature Date

Data Entry Comments:

uploaded 2-17-99
JL Wauell
28270FEB.CSV

validated 2/24/99 Jm Faye

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 61N21-B LCS Date: 02/14/1999 14:08:41
 Data File : C:\DX\DATA\99021401.D45
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 45 Detector: CDM-1
 Analyst : *Jeffrey Purinton* Column: AG4A/AS4A anion column

02-14-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	91	335	2	
2	0.99	fluoride	64.484108,93	1853	9798	2	-0.34
3	1.42	chloride	82.172103,23	2403	10119	1	-0.70
4	1.67	nitrite	496.01785,59	8616	41997	1	-2.14
5	2.42	bromide	570.05097,28	5993	32911	1	-5.47
6	2.73	nitrate	593.097100,19	7233	46403	1	-1.21
7	3.99	phosphate	544.46699,36	1429	17989	1	0.84
8	5.07	sulfate	627.45198,35	4532	60722	1	1.13
9	6.56	oxalate	601.237108,72	2480	41004	1	1.65
Totals			3578.974	34630	261278		

File: 99021401.D45 Sample: 61N21-B LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1678 TO 1685.

```

=====
Sample Name: 60N21-A CCV                      Date: 02/14/1999 14:20:09
Data File  : C:\DX\DATA\99021401.D46
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 46           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          101    3000  5Hz   0.00 10.00          30
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	102	375	2	
2	0.99	fluoride	62.072	1838	9426	2	-0.34
3	1.43	chloride	89.515	2625	11037	1	-0.23
4	1.68	nitrite	565.955	9493	47937	1	-1.75
5	2.43	bromide	608.779	6479	35185	1	-5.21
6	2.73	nitrate	670.348	8297	52543	1	-0.97
7	3.99	phosphate	630.083	1690	20883	1	0.84
8	5.07	sulfate	673.398	4593	65241	1	1.13
9	6.56	oxalate	587.262	2299	40044	1	1.65
Totals			3887.411	37415	282671		

File: 99021401.D46 Sample: 60N21-A CCV

```

=====
Sample Name: BLANK                      Date: 02/14/1999 13:57:28
Data File  : C:\DX\DATA\99021401.D44
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 44    Detector: CDM-1
Analyst    :                          Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      1      3000  5Hz   0.00 10.00      30
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.18		0.000	38	355	1	
2	0.83		0.000	51	245	1	
3	1.35	chloride	0.083	114	1021	1	-5.36
Totals			0.083	203	1621		

```

=====
Sample Name: PREP BLANK                      Date: 02/14/1999 14:31:24
Data File  : C:\DX\DATA\99021401.D47
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 47         Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      1      3000  5Hz  0.00 10.00      30
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.27		0.000	57	444	1	
2	1.44	chloride	0.013	39	156	1	0.70
Totals			0.013	96	600		

```

=====
Sample Name: S98T002849 SAM                      Date: 02/14/1999 16:14:05
Data File  : F:\DATA\99021401.D56
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 56              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00         30
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	66	270	2	
2	0.99	fluoride	9.955	300	1493	2	-0.34
3	1.41	chloride	8.972	234	1092	1	-1.17
4	1.67	nitrite	98.676	1696	8336	1	-2.53
5	2.69	nitrate	1483.774	17467	118644	1	0.00
6	4.00	phosphate	86.412	243	2809	1	1.01
7	5.12	sulfate	19.787	153	1888	1	2.20
8	6.61	oxalate	29.912	121	2024	1	2.48
Totals			1737.488	20281	136555		

File: 99021401.D56 Sample: S98T002849 SAM

Sample Name: S98T002849	DUP	Date: 02/14/1999 16:24:51
Data File : F:\DATA\99021401.D57		
Method : C:\DX\METHOD\KIT.MET		
ACI Address: 1	System: 1	Inject#: 57
Analyst :	Column: AG4A/AS4A anion column	Detector: CDM-1

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	64	253	2	
2	0.99	fluoride	3.766	121	564	2	-0.34
3	1.42	chloride	8.399	216	1022	1	-0.70
4	1.67	nitrite	91.482	1563	7728	1	-2.53
5	2.69	nitrate	1782.999	20845	143666	1	0.00
6	4.01	phosphate	26.313	83	854	1	1.18
7	5.12	sulfate	13.279	106	1267	1	2.20
8	6.61	oxalate	22.304	100	1509	1	2.48
Totals			1948.542	23097	156862		

File: 99021401.D57 Sample: S98T002849 DUP

```

=====
Sample Name: S98T002931 SAM                      Date: 02/14/1999 15:51:43
Data File  : F:\DATA\99021401.D54
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 54              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          101    3000   5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	73	295	2	
2	0.98	fluoride	4.057	144	607	2	-1.01
3	1.42	chloride	3.636	101	442	1	-0.70
4	1.67	nitrite	43.637	739	3685	1	-2.53
5	2.67	nitrate	2744.340	31781	226930	1	0.00
6	4.01	phosphate	22.462	69	729	1	1.35
7	5.12	sulfate	17.046	133	1626	1	2.20
8	6.61	oxalate	20.954	93	1418	1	2.48

Totals 2856.132 33133 235732

File: 99021401.D54 Sample: S98T002931 SAM

```

=====
Sample Name: S98T002931  DUP                               Date: 02/14/1999 16:02:41
Data File  : F:\DATA\99021401.D55
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 55                      Detector:CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	76	296	2	
2	0.99	fluoride	4.134	150	619	2	-0.34
3	1.41	chloride	4.147	103	504	1	-1.17
4	1.67	nitrite	39.660	693	3349	1	-2.53
5	2.66	nitrate	3011.739	34398	250940	1	0.00
6	4.00	phosphate	24.711	82	802	1	1.01
7	5.12	sulfate	15.491	122	1478	1	2.20
8	6.61	oxalate	17.874	81	1209	1	2.48
Totals			3117.755	35705	259197		

File: 99021401.D55 Sample: S98T002931 DUP

LABCORE Completed Worklist Report for Worklist# 28318

Analyst: kjt

Instrument: IC40S1

Book#: 66N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 for @IC-01

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	@IC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	@IC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	@IC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	@IC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	@IC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	@IC-QC P04	QC	1	<1.20e-1		ug/mL
1	CCB	0	@IC-QC S04	QC	1	<1.38e-1		ug/mL
1	CCB	0	@IC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	LCS-INST	0	@IC-QC F	QC	5.84e1	5.74e+01	98.288 % Recovery	
2	LCS-INST	0	@IC-QC CL	QC	8.06e1	8.28e+01	102.730 % Recovery	
2	LCS-INST	0	@IC-QC NO2	QC	5.50e2	5.84e+02	106.182 % Recovery	
2	LCS-INST	0	@IC-QC BR	QC	5.76e2	5.83e+02	101.215 % Recovery	
2	LCS-INST	0	@IC-QC NO3	QC	5.84e2	5.46e+02	93.493 % Recovery	
2	LCS-INST	0	@IC-QC P04	QC	5.44e2	5.56e+02	102.206 % Recovery	
2	LCS-INST	0	@IC-QC S04	QC	6.42e2	6.54e+02	101.869 % Recovery	
2	LCS-INST	0	@IC-QC OXALATE2	QC	5.30e2	5.39e+02	101.698 % Recovery	
3	CCV	0	@IC-QC F	QC	6.40e1	6.33e+01	98.906 % Recovery	
3	CCV	0	@IC-QC CL	QC	8.95e1	8.93e+01	99.777 % Recovery	
3	CCV	0	@IC-QC NO2	QC	5.23e2	4.96e+02	94.837 % Recovery	
3	CCV	0	@IC-QC BR	QC	6.30e2	6.07e+02	96.349 % Recovery	
3	CCV	0	@IC-QC NO3	QC	7.00e2	6.76e+02	96.571 % Recovery	
3	CCV	0	@IC-QC P04	QC	6.35e2	6.07e+02	95.591 % Recovery	
3	CCV	0	@IC-QC S04	QC	7.00e2	6.69e+02	95.571 % Recovery	
3	CCV	0	@IC-QC OXALATE2	QC	5.30e2	5.10e+02	96.226 % Recovery	
4	BLNK-PREP	0	@IC-01 F-02	SOLID	1	<1.20e-2		ug/g
4	BLNK-PREP	0	@IC-01 CL-02	SOLID	1	5.40e-02	0.054	ug/g
4	BLNK-PREP	0	@IC-01 NO2-02	SOLID	1	<1.08e-1		ug/g
4	BLNK-PREP	0	@IC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4	BLNK-PREP	0	@IC-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4	BLNK-PREP	0	@IC-01 P04-02	SOLID	1	<1.20e-1		ug/g
4	BLNK-PREP	0	@IC-01 S04-02	SOLID	1	<1.38e-1		ug/g
4	BLNK-PREP	0	@IC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5	SAMPLE	S98T002848	0 W @IC-01 F-02	SOLID	N/A	6.735e+02	237.200	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 CL-02	SOLID	N/A	1.130e+03	336.100	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 NO2-02	SOLID	N/A	1.425e+04	2135.000	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 BR-02	SOLID	N/A	2.472e+03	2471.000	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 NO3-02	SOLID	N/A	5.319e+05	2748.000	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 P04-02	SOLID	N/A	5.494e+03	2372.000	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 S04-02	SOLID	N/A	2.877e+03	2729.000	ug/g
5	SAMPLE	S98T002848	0 W @IC-01 OXALATE2	SOLID	N/A	3.044e+03	2077.000	ug/g
6	DUP	S98T002848	0 W @IC-01 F-02	SOLID	6.74e+02	1.69e+03	85.956	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28318

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 DUP	S98T002848	0 W	@IC-01	CL-02	SOLID	1.13e+03	1.23e+03	8.475 RPD
6 DUP	S98T002848	0 W	@IC-01	NO2-02	SOLID	1.42e+04	1.52e+04	6.803 RPD
6 DUP	S98T002848	0 W	@IC-01	BR-02	SOLID	<2.47e3		RPD
6 DUP	S98T002848	0 W	@IC-01	NO3-02	SOLID	5.32e+05	5.04e+05	5.405 RPD
6 DUP	S98T002848	0 W	@IC-01	PO4-02	SOLID	5.49e+03	8.15e+03	39.003 RPD
6 DUP	S98T002848	0 W	@IC-01	SO4-02	SOLID	2.68e+03	3.58e+03	21.672 RPD
6 DUP	S98T002848	0 W	@IC-01	OxALATE2	SOLID	3.04e+03	3.35e+03	9.703 RPD
7 SPK	S98T002848	0 W	@IC-01	F-02	SOLID	5.84e1	4.20e+01	71.918 % Recovery
7 SPK	S98T002848	0 W	@IC-01	CL-02	SOLID	8.06e1	5.78e+01	71.712 % Recovery
7 SPK	S98T002848	0 W	@IC-01	NO2-02	SOLID	5.50e2	3.90e+02	70.909 % Recovery
7 SPK	S98T002848	0 W	@IC-01	BR-02	SOLID	5.76e2	3.92e+02	68.056 % Recovery
7 SPK	S98T002848	0 W	@IC-01	NO3-02	SOLID	5.84e2	4.23e+02	72.432 % Recovery
7 SPK	S98T002848	0 W	@IC-01	PO4-02	SOLID	5.44e2	3.75e+02	68.934 % Recovery
7 SPK	S98T002848	0 W	@IC-01	SO4-02	SOLID	6.42e2	4.40e+02	68.536 % Recovery
7 SPK	S98T002848	0 W	@IC-01	OxALATE2	SOLID	5.30e2	3.71e+02	70.000 % Recovery
8 BLNK-PREP		0	@IC-01	F-02	SOLID	1	<1.20e-2	ug/g
8 BLNK-PREP		0	@IC-01	CL-02	SOLID	1	4.20e-02	0.042 ug/g
8 BLNK-PREP		0	@IC-01	NO2-02	SOLID	1	<1.08e-1	ug/g
8 BLNK-PREP		0	@IC-01	BR-02	SOLID	1	<1.25e-1	ug/g
8 BLNK-PREP		0	@IC-01	NO3-02	SOLID	1	<1.39e-1	ug/g
8 BLNK-PREP		0	@IC-01	PO4-02	SOLID	1	<1.20e-1	ug/g
8 BLNK-PREP		0	@IC-01	SO4-02	SOLID	1	<1.38e-1	ug/g
8 BLNK-PREP		0	@IC-01	OxALATE2	SOLID	1	<1.05e-1	ug/g
9 SAMPLE	S98T002613	0 W	@IC-01	NO3-02	SOLID	N/A	7.130e+05	2839.000 ug/g
10 DUP	S98T002613	0 W	@IC-01	F-02	SOLID	?	5.75e+02	RPD
10 DUP	S98T002613	0 W	@IC-01	CL-02	SOLID	?	<3.41e2	RPD
10 DUP	S98T002613	0 W	@IC-01	NO2-02	SOLID	?	3.15e+03	RPD
10 DUP	S98T002613	0 W	@IC-01	BR-02	SOLID	?	<2.50e3	RPD
10 DUP	S98T002613	0 W	@IC-01	NO3-02	SOLID	7.13e+05	6.47e+05	9.706 RPD
10 DUP	S98T002613	0 W	@IC-01	PO4-02	SOLID	?	3.04e+03	RPD
10 DUP	S98T002613	0 W	@IC-01	SO4-02	SOLID	?	<2.76e3	RPD
10 DUP	S98T002613	0 W	@IC-01	OxALATE2	SOLID	?	2.34e+03	RPD

Final page for worklist# 28318

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

James M. L...
Reviewer Signature _____ Date 3/12/99

LABCORE Data Entry Template for Worklist# 28318

Analyst: KJT Instrument: IC 4051 Book# 661218Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 for @IC-01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002848 0 W		@IC-01	SOLID	98000437	BY-105 (2)
Analytes Requested: BR-02, CL-02, F-02, NO2-02, NO3-02, OXALATE2, PO4-02, SO4-02						
6 DUP	S98T002848 0 W		@IC-01	SOLID		
7 SPK	S98T002848 0 W		@IC-01	SOLID		
8 BLNK-PREP			@IC-01	SOLID		
9 SAMPLE	S98T002613 0 W		@IC-01	SOLID	98000436	BY-105 (2)
Analytes Requested: NO3-02						
10 DUP	S98T002613 0 W		@IC-01	SOLID		

Final page for worklist # 28318

See original

Signature Date

Signature Date

Data Entry Comments:

28318.m2.csv reviewed & reuploaded to correct standard entry error. 3/12/99pm

Validated 3/12/99 RPDs & Spk pool for S98T002848 on this

result.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

03/10/99 08:37

ws2

Page: 1

LABCORE Data Entry Template for Worklist# 28318

Analyst: KST Instrument: IC 4051 Book# 66W21BMethod: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 for @IC-01

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC-01	SOLID		
5	SAMPLE	S98T002848	0	W	@IC-01	SOLID	98000437	BY-105 (2)
Analytes Requested: BR-02, CL-02, F-02, NO2-02, NO3-02, OXALATE2, PO4-02, SO4-02								
6	DUP	S98T002848	0	W	@IC-01	SOLID		
7	SPK	S98T002848	0	W	@IC-01	SOLID		

Final page for worklist # 28318

KS Pomele 3-10-99
 Signature Date

_____
 Signature Date

Data Entry Comments:

Continue with (28265) ³¹⁶⁻⁹⁹ - Make new
 combined worklist before data entry

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1689

Data Reprocessed On 03/12/1999 10:33:11

Sample Name: INSTR BLANK Date: 03/10/1999 08:50:04
 Data File : F:\DATA\99031001.d53
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 53 Detector: CDM-1
 Analyst : *Janet for KJT* Column: AG4A/AS4A anion column

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 1 2895 5Hz 0.00 9.65 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.37		0.000	12	64		1
Totals			0.000	12	64		

File: 99031001.d53 Sample: INSTR BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1680 TO 1689.

Data Reprocessed On 03/12/1999 10:33:11

```

=====
Sample Name: 66N21-B LCS                      Date: 03/10/1999 09:29:13
Data File  : F:\DATA\99031001.d56
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 56           Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	81	382	1	
2	1.05	fluoride	57.355	1662	9155	1	0.64
3	1.65	chloride	82.753	1986	9964	1	0.82
4	2.00	nitrite	584.399	8142	49201	1	0.00
5	3.08	bromide	582.892	4330	32193	1	0.00
6	3.53	nitrate	545.907	4527	41780	1	0.09
7	5.07	phosphate	555.992	1215	20020	1	-1.04
8	6.83	sulfate	654.258	3613	60228	1	-0.05
9	9.07	oxalate	539.180	1741	39019	1	-0.58
Totals			3602.736	27296	261942		

File: 99031001.d56 Sample: 66N21-B LCS

Data Reprocessed On 03/12/1999 10:33:11

Sample Name: 60N21-B CCV	Date: 03/10/1999 09:43:31
Data File : F:\DATA\99031001.d57	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 57	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	88	415	2	
2	1.05	fluoride	63.251	1757	10113	3	0.64
3	1.35		0.000	64	490	4	
4	1.64	chloride	89.336	2111	10770	1	0.41
5	2.00	nitrite	496.041	6782	41695	1	0.00
6	3.07	bromide	607.451	4543	33578	1	-0.22
7	3.52	nitrate	675.683	5573	51889	1	-0.28
8	5.07	phosphate	607.056	1440	21891	1	-1.04
9	6.83	sulfate	668.575	3730	61569	1	-0.05
10	9.07	oxalate	509.657	1656	36867	1	-0.58
Totals			3717.050	27746	269276		

File: 99031001.d57 Sample: 60N21-B CCV

Data Reprocessed On 03/12/1999 10:33:12

```

=====
Sample Name: PREP-BNLK for 99031001.d63 Date: 03/10/1999 13:54:40
Data File  : F:\DATA\99031001.d63
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 63      Detector: CDM-1
Analyst    :      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.29		0.000	51	432	1	
2	1.66	chloride	0.054	67	650	1	1.63
Totals			0.054	118	1082		

Data Reprocessed On 03/12/1999 10:33:13

```

=====
Sample Name: S98T002848                      Date: 03/10/1999 14:05:30
Data File  : F:\DATA\99031001.d64
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 64          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101      3000  5Hz  0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	77	400	1	
2	1.05	fluoride	3.440	126	541	1	1.28
3	1.65	chloride	5.770	142	685	1	0.82
4	2.00	nitrite	72.810	1016	6074	1	0.00
5	3.44	nitrate	2716.789	21289	221299	1	-2.55
6	5.12	phosphate	28.065	71	996	1	0.00
7	6.88	sulfate	14.696	80	1330	1	0.73
8	9.12	oxalate	15.550	58	1117	1	0.00
Totals			2857.120	22858	232442		

File: 99031001.d64 Sample: S98T002848

Data Reprocessed On 03/12/1999 10:33:13

```

=====
Sample Name: S98T002848 DUP                      Date: 03/10/1999 14:17:21
Data File  : F:\DATA\99031001.d65
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 65              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000  5Hz   0.00 10.00        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	76	379	2	
2	1.05	fluoride	8.630	238	1360	2	1.28
3	1.65	chloride	6.282	150	746	1	0.82
4	2.00	nitrite	77.805	1114	6491	1	0.00
5	3.45	nitrate	2575.084	20086	208823	1	-2.36
6	5.12	phosphate	41.607	115	1476	1	0.00
7	6.88	sulfate	18.277	104	1655	1	0.73
8	9.12	oxalate	17.135	55	1231	1	0.00
Totals			2744.819	21938	222160		

File: 99031001.d65 Sample: S98T002848 DUP

Sample Name: S98T002848 SPIKE Date: 03/10/1999 14:29:38
 Data File : F:\DATA\99031001.d66
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 66 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	144	705	2	
2	1.05	fluoride	45.418	1251	7226	2	0.64
3	1.64	chloride	63.557	1463	7625	1	0.41
4	2.00	nitrite	462.733	6463	38872	1	0.00
5	3.07	bromide	391.971	3069	21509	1	-0.22
6	3.42	nitrate	3139.376	24183	259247	1	-3.12
7	5.07	phosphate	402.868	913	14442	1	-1.04
8	6.83	sulfate	454.818	2485	41646	1	-0.05
9	9.07	oxalate	386.571	1242	27915	1	-0.58
Totals			5347.311	41211	419187		

File: 99031001.d66 Sample: S98T002848 SPIKE

Data Reprocessed On 03/12/1999 10:33:11

Sample Name: PREP BLANK 2613	Date: 03/10/1999 09:55:37
Data File : F:\DATA\99031001.d58	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 58	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.29		0.000	50	396	1	
2	1.65	chloride	0.042	48	506	1	1.22
Totals			0.042	98	903		

File: 99031001.d58 Sample: PREP BLANK 2613

Data Reprocessed On 03/12/1999 10:33:12

```

=====
Sample Name: S98T002613                      Date: 03/10/1999 10:02:05
Data File  : F:\DATA\99031001.d59
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 59          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    1541  5Hz   0.00  5.13        30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	92	451	2	
2	1.05	fluoride	2.987	88	470	2	0.64
3	1.65	chloride	1.416	38	168	1	0.82
4	1.99	nitrite	15.725	227	1310	1	-0.33
5	3.41	nitrate	3525.972	27488	295003	1	-3.49
Totals			3546.100	27934	297403		

File: 99031001.d59 Sample: S98T002613

```

=====
Sample Name: S98T002613 DUP                      Date: 03/10/1999 10:16:09
Data File  : F:\DATA\99031001.d60
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 60              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    3000   5Hz   0.00  10.00          30
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.87		0.000	88	493	2	
2	1.05	fluoride	2.901	82	456	2	1.28
3	1.64	chloride	1.243	33	147	1	0.41
4	1.99	nitrite	15.876	224	1323	1	-0.33
5	3.41	nitrate	3260.399	25357	270329	1	-3.31
6	5.12	phosphate	15.330	45	544	1	0.00
7	6.83	sulfate	9.121	46	825	1	-0.05
8	9.12	oxalate	11.777	39	846	1	0.00
Totals			3316.646	25915	274964		

File: 99031001.d60 Sample: S98T002613 DUP

LABCORE Completed Worklist Report for Worklist# 28341

Analyst: krm

Instrument: IC40S1

Book#: 60N21B

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	QIC-QC F	QC	1	<1.20e-2		ug/mL
1	CCB	0	QIC-QC CL	QC	1	<1.70e-2		ug/mL
1	CCB	0	QIC-QC NO2	QC	1	<1.08e-1		ug/mL
1	CCB	0	QIC-QC BR	QC	1	<1.25e-1		ug/mL
1	CCB	0	QIC-QC NO3	QC	1	<1.39e-1		ug/mL
1	CCB	0	QIC-QC P04	QC	1	<1.20e-1		ug/mL
1	CCB	0	QIC-QC S04	QC	1	<1.38e-1		ug/mL
1	CCB	0	QIC-QC OXALATE2	QC	1	<1.05e-1		ug/mL
2	LCS-INST	0	QIC-QC F	QC	5.92e1	6.01e+01	101.520 %	Recovery
2	LCS-INST	0	QIC-QC CL	QC	7.96e1	7.98e+01	100.251 %	Recovery
2	LCS-INST	0	QIC-QC NO2	QC	5.53e2	4.81e+02	86.980 %	Recovery
2	LCS-INST	0	QIC-QC BR	QC	5.86e2	5.60e+02	95.563 %	Recovery
2	LCS-INST	0	QIC-QC NO3	QC	5.92e2	6.23e+02	105.236 %	Recovery
2	LCS-INST	0	QIC-QC P04	QC	5.48e2	5.20e+02	94.891 %	Recovery
2	LCS-INST	0	QIC-QC S04	QC	6.38e2	6.15e+02	96.395 %	Recovery
2	LCS-INST	0	QIC-QC OXALATE2	QC	5.33e2	5.85e+02	109.756 %	Recovery
3	CCV	0	QIC-QC F	QC	6.40e1	6.29e+01	98.281 %	Recovery
3	CCV	0	QIC-QC CL	QC	8.95e1	9.32e+01	104.134 %	Recovery
3	CCV	0	QIC-QC NO2	QC	5.23e2	5.19e+02	103.059 %	Recovery
3	CCV	0	QIC-QC BR	QC	6.30e2	6.18e+02	98.095 %	Recovery
3	CCV	0	QIC-QC NO3	QC	7.00e2	6.80e+02	97.143 %	Recovery
3	CCV	0	QIC-QC P04	QC	6.35e2	6.23e+02	98.110 %	Recovery
3	CCV	0	QIC-QC S04	QC	7.00e2	6.89e+02	98.429 %	Recovery
3	CCV	0	QIC-QC OXALATE2	QC	5.30e2	5.83e+02	110.000 %	Recovery
4	BLNK-PREP	0	QIC-01 F-02	SOLID	1	<1.20e-2		ug/g
4	BLNK-PREP	0	QIC-01 CL-02	SOLID	1	<1.70e-2		ug/g
4	BLNK-PREP	0	QIC-01 NO2-02	SOLID	1	<1.08e-1		ug/g
4	BLNK-PREP	0	QIC-01 BR-02	SOLID	1	<1.25e-1		ug/g
4	BLNK-PREP	0	QIC-01 NO3-02	SOLID	1	<1.39e-1		ug/g
4	BLNK-PREP	0	QIC-01 P04-02	SOLID	1	<1.20e-1		ug/g
4	BLNK-PREP	0	QIC-01 S04-02	SOLID	1	<1.38e-1		ug/g
4	BLNK-PREP	0	QIC-01 OXALATE2	SOLID	1	<1.05e-1		ug/g
5	SAMPLE	S98T002714	0 W QIC-01 F-02	SOLID	N/A	1.319e+04	120.000	ug/g
5	SAMPLE	S98T002714	0 W QIC-01 CL-02	SOLID	N/A	1.478e+03	170.000	ug/g
5	SAMPLE	S98T002714	0 W QIC-01 NO3-02	SOLID	N/A	6.961e+04	1390.000	ug/g
5	SAMPLE	S98T002714	0 W QIC-01 P04-02	SOLID	N/A	6.480e+04	1200.000	ug/g
6	DUP	S98T002714	0 W QIC-01 F-02	SOLID	1.32e+04	7.54e+03	54.581	RPD
6	DUP	S98T002714	0 W QIC-01 CL-02	SOLID	1.48e+03	1.23e+03	18.450	RPD
6	DUP	S98T002714	0 W QIC-01 NO2-02	SOLID	?	1.81e+04		RPD
6	DUP	S98T002714	0 W QIC-01 BR-02	SOLID	?	<1.23e3		RPD
6	DUP	S98T002714	0 W QIC-01 NO3-02	SOLID	6.96e+04	5.17e+04	29.514	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

1700

LABCORE Completed Worklist Report for Worklist# 28341

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
6	DUP	S98T002714	0	W	@IC-01	P04-02	SOLID	6.48e+04	1.28e+04	134.021 RPD
6	DUP	S98T002714	0	W	@IC-01	S04-02	SOLID	?	2.93e+04	RPD
6	DUP	S98T002714	0	W	@IC-01	OxALATE2	SOLID	?	1.32e+04	RPD

Final page for worklist# 28341

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 3/12/99

02/22/99 13:15
ws2

Page: 1

LABCORE Data Entry Template for Worklist# 28341

Analyst: KRM Instrument: IC 4051 Book# 61 N21D
60N21B

Method: LA-533-105 Rev/Mod _____

Worklist Comment: BY105 FOR @IC-01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002714 0 W	@IC-01	SOLID	98000436 BY-105 (2)		
	Analytes Requested: CL-02		, F-02	, NO3-02 , PO4-02		
6 DUP	S98T002714 0 W	@IC-01	SOLID			

Final page for worklist # 28341

Signature

Date

Signature

Date

Data Entry Comments:

up by lab 8-3-99

J. Warrall

28341 FEB. CSV

Validated 3/9/99 msh

RPD's still high on this

run, probably still non-homogeneous

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 02/27/1999 19:49:08
 Data File : C:\DX\DATA\99022711.D02
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 2 Detector: CDM-1
 Analyst: *[Signature]* Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.41	chloride	0.012	24	146	1	-1.63
2	6.45	oxalate	0.071	30	482	1	0.00
3	7.41		0.000	14	264	1	
Totals			0.082	68	891		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1703 TO 1708.

Sample Name: LCS 61N21D Date: 02/27/1999 20:00:46
 Data File : C:\DX\DATA\99022711.D03
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 3 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	76	282	2	
2	0.99	fluoride	60.142	1747	9129	2	-0.34
3	1.43	chloride	79.835	2322	9828	1	-0.23
4	1.68	nitrite	481.074	8156	40728	1	-1.75
5	2.43	bromide	560.288	5977	32338	1	-4.95
6	2.74	nitrate	623.251	7737	48797	1	-0.72
7	3.97	phosphate	519.930	1419	17163	1	0.34
8	5.07	sulfate	614.540	5014	59454	1	1.13
9	6.56	oxalate	585.421	2640	39917	1	1.65
Totals			3524.481	35088	257635		

File: 99022711.D03 Sample: LCS 61N21D

Sample Name: CCV 60N21B Date: 02/27/1999 20:14:47
 Data File : C:\DX\DATA\99022711.D04
 Method : C:\DX\METHOD\KIT.MET
 ACI Address: 1 System: 1 Inject#: 4 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.84		0.000	73	261	2	
2	0.98	fluoride	62.920	1853	9557	2	-1.01
3	1.41	chloride	93.181	2643	11496	1	-1.17
4	1.66	nitrite	539.156	9120	45660	1	-2.92
5	2.41	bromide	618.083	6700	35732	1	-5.73
6	2.71	nitrate	680.409	8458	53344	1	-1.69
7	3.95	phosphate	622.573	1721	20629	1	-0.34
8	5.01	sulfate	689.430	4573	66820	1	0.07
9	6.51	oxalate	582.735	2429	39732	1	0.83
Totals			3888.486	37571	283231		

File: 99022711.D04 Sample: CCV 60N21B

```

=====
Sample Name: PREP BLK                      Date: 02/27/1999 20:31:26
Data File  : C:\DX\DATA\99022711.D05
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 5        Detector: CDM-1
Analyst    :                            Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.25		0.000	61	548	1	
2	1.43	chloride	0.014	53	172	1	-0.23
3	1.69	nitrite	0.009	18	80	1	-1.36
Totals			0.023	133	799		

Data Reprocessed On 03/03/1999 11:46:03

```

=====
Sample Name: S98T002714                      Date: 02/27/1999 21:41:11
Data File  : F:\DATA\99022711.D06
Method     : C:\DX\METHOD\KIT.MET
ACI Address: 1 System: 1 Inject#: 6           Detector:CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51    3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	97	340	2	
2	0.98	fluoride	67.277	3882	20603	2	-1.01
3	1.41	chloride	7.535	413	1818	2	-1.17
4	1.67	nitrite	108.392	3504	18146	1	-2.53
5	2.71	nitrate	354.955	8669	55141	1	0.00
6	3.95	phosphate	330.413	1806	21707	1	-0.34
7	5.07	sulfate	138.311	2318	26285	1	1.13
8	6.56	oxalate	72.592	659	9743	1	1.65
Totals			1079.475	21347	153783		

Data Reprocessed On 03/03/1999 11:46:57

Sample Name: S98T002714 DUP	Date: 02/27/1999 21:56:32
Data File : F:\DATA\99022711.D07	
Method : C:\DX\METHOD\KIT.MET	
ACI Address: 1 System: 1 Inject#: 7	Detector: CDM-1
Analyst :	Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	51	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.85		0.000	134	499	2	
2	0.98	fluoride	38.978	2260	11766	2	-1.01
3	1.41	chloride	6.347	358	1531	1	-1.17
4	1.67	nitrite	93.509	3092	15652	1	-2.53
5	2.72	nitrate	267.324	6512	41357	1	0.00
6	3.95	phosphate	66.094	365	4262	1	-0.17
7	5.07	sulfate	151.617	2555	28832	1	1.13
8	6.56	oxalate	68.318	631	9169	1	1.65
Totals			692.188	15907	113067		

File: 99022711.D07 Sample: S98T002714 DUP

LABCORE Completed Worklist Report for Worklist# 28346

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 rerun for RPD and spk failure. Try 21df. jmf

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0 @IC4G-QC F*4	QC		1	<1.10e-2		ug/mL
1 CCB	0 @IC4G-QC ACETATE2	QC		1	2.80e-02	0.028	ug/mL
1 CCB	0 @IC4G-QC FORMATE2	QC		1	<2.10e-2		ug/mL
1 CCB	0 @IC4G-QC GLYCOLTI	QC		1	<2.00e-2		ug/mL
2 LCS-INST	0 @IC4G-QC F*4	QC		6.10e1	6.31e+01	103.443 % Recovery	
2 LCS-INST	0 @IC4G-QC ACETATE2	QC		1.16e2	1.15e+02	99.138 % Recovery	
2 LCS-INST	0 @IC4G-QC FORMATE2	QC		1.19e2	1.24e+02	104.202 % Recovery	
2 LCS-INST	0 @IC4G-QC GLYCOLTI	QC		1.08e2	1.03e+02	95.370 % Recovery	
3 CCV	0 @IC4G-QC F*4	QC		6.29e1	5.63e+01	89.507 % Recovery	
3 CCV	0 @IC4G-QC ACETATE2	QC		1.20e2	1.14e+02	95.000 % Recovery	
3 CCV	0 @IC4G-QC FORMATE2	QC		1.20e2	1.21e+02	100.833 % Recovery	
3 CCV	0 @IC4G-QC GLYCOLTI	QC		1.03e2	9.47e+01	91.942 % Recovery	
4 BLNK-PREP	0 @IC4G-01 F*4-01	SOLID		1	<3.10e-2		ug/g
4 BLNK-PREP	0 @IC4G-01 ACETATE2	SOLID		1	2.90e-02	0.029	ug/g
4 BLNK-PREP	0 @IC4G-01 FORMATE2	SOLID		1	<6.20e-2		ug/g
4 BLNK-PREP	0 @IC4G-01 GLYCOLTI	SOLID		1	<5.60e-2		ug/g
5 SAMPLE	S98T002849 0 W @IC4G-01 ACETATE2	SOLID		N/A	9.06e+02	84.320	ug/g
5 SAMPLE	S98T002849 0 W @IC4G-01 FORMATE2	SOLID		N/A	5.28e+02	261.500	ug/g
6 DUP	S98T002849 0 W @IC4G-01 F*4-01	SOLID		7	9.07e+02		RPD
6 DUP	S98T002849 0 W @IC4G-01 ACETATE2	SOLID		9.06e+02	7.88e+02	13.932	RPD
6 DUP	S98T002849 0 W @IC4G-01 FORMATE2	SOLID		5.28e+02	4.32e+02	20.000	RPD
6 DUP	S98T002849 0 W @IC4G-01 GLYCOLTI	SOLID		7	<2.33e2		RPD
8 DUP	S98T002932 0 W @IC4G-01 F*4-01	SOLID		2.08e+03	2.08e+03	0.000	RPD
8 DUP	S98T002932 0 W @IC4G-01 ACETATE2	SOLID		7	2.78e+02		RPD
8 DUP	S98T002932 0 W @IC4G-01 FORMATE2	SOLID		7	<2.64e2		RPD
8 DUP	S98T002932 0 W @IC4G-01 GLYCOLTI	SOLID		7	<2.38e2		RPD
9 SPK	S98T002932 0 W @IC4G-01 F*4-01	SOLID		6.10e1	6.86e+01	112.459 % Recovery	
9 SPK	S98T002932 0 W @IC4G-01 ACETATE2	SOLID		1.16e2	0.00e+00	0.000 % Recovery	
9 SPK	S98T002932 0 W @IC4G-01 FORMATE2	SOLID		1.19e2	0.00e+00	0.000 % Recovery	
9 SPK	S98T002932 0 W @IC4G-01 GLYCOLTI	SOLID		1.08e2	0.00e+00	0.000 % Recovery	

Final page for worklist# 28346

Analyst Signature

Date

Analyst Signature

Date

James L. Lyle
Reviewer Signature

3/2/99
Date

02/24/99 08:32

WS2

Page: 1

LABCORE Data Entry Template for Worklist# 28346

Analyst: ADP Instrument: IC _____ Book# 59N21-A LCS
58N21-A CCV
 Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 rerun for RPD and spk failure. Try 21df. jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002849 0 W		@IC4G-01	SOLID	98000437	BY-105 (2)
Analytes Requested: ACETATE2, FORMATE2						
6 DUP	S98T002849 0 W		@IC4G-01	SOLID		
7 SAMPLE	S98T002932 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
Analytes Requested: F*4-01						
8 DUP	S98T002932 0 W		@IC4G-01	SOLID		
9 SPK	S98T002932 0 W		@IC4G-01	SOLID		

Final page for worklist # 28346

Arthur Purinton 02-25-99
 Signature Date

Signature Date

Data Entry Comments:

up loaded 2-26-99
John Wall
28346 FEB. CCV
1710 F CCV failed so S98T002932 was rejected & will need to be rerun. Rest Validated
3/2/99 9M jmf
reanalyzed on w/ 28541. f 3/24/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: 59N21-A LCS Date: 02/25/1999 14:41:40
 Data File : C:\DX\DATA\99022521.D20
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 20 Detector: CDM-2
 Analyst : *Anthony Panto* Column: AG14A-SC, AS14A-SC, SRS

02-25-98

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	20372	706642	1	
25	7.20	fluoride	63.104	10345	980664	2	0.00
26	7.89	glycolate	102.815	520	535817	2	0.00
27	8.37	acetate	114.988	9913	546608	2	0.00
28	9.07	formate	123.501	10378	039962	2	0.00
Totals			404.408	4123423	68349210		

File: 99022521.D20 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1711 TO 1720

1711

Sample Name: 58N21-A CCV Date: 02/25/1999 09:45:01
 Data File : C:\DX\DATA\99022521.D17
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 17 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4713	5Hz	0.00	15.71	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	5166	108666	1	
3	2.25		0.000	3274	31584	2	
7	2.93		0.000	2454	22952	2	
8	3.08		0.000	1994	18491	2	
9	3.27		0.000	1931	31359	2	
18	5.52		0.000	3801	68657	2	
23	7.15	fluoride	56.255	1764643	25536692	2	0.00
24	7.81	glycolate	94.702	470431	8560341	2	0.00
25	8.29	acetate	113.558	540086	10382179	2	0.00
26	8.99	formate	121.165	97978322	18606356	2	0.00
Totals			385.679	3772101	63367276		

File: 99022521.D17 Sample: 58N21-A CCV

1712

```

=====
Sample Name: INSTR BLANK                      Date: 02/25/1999 08:50:35
Data File  : C:\DX\DATA\99022521.D14
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 14          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1  5017 5Hz  0.00 16.72      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.50		0.000	90018	2388308	1	
2	1.67		0.000	212750	2078440	1	
3	3.05		0.000	1091239	56130660	1	
6	7.49		0.000	2000	23600	1	
7	8.37	acetate	0.028	850	27600	1	0.00
8	9.36		0.000	3011	16400	1	
Totals			0.028	1399867	60665008		

File: 99022521.D14 Sample: INSTR BLANK

1713

Sample Name: BLANK PREP

Date: 02/25/1999 15:03:32

Data File : C:\DX\DATA\99022521.D21

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 21

Detector: CDM-2

Analyst :

Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject

External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	13287	631388	1	
5	2.88		0.000	2702	25708	2	
6	3.01		0.000	3814	28538	2	
8	3.21		0.000	3282	21026	2	
9	3.36		0.000	2894	16375	2	
10	3.49		0.000	2786	26154	2	
11	3.65		0.000	2796	26121	2	
12	3.96		0.000	1438	20054	2	
13	4.43		0.000	2111	20040	2	
26	7.23	fluoride	0.024	2251	45000	2	0.00
29	8.27	acetate	0.029	1817	36747	2	0.00
Totals			0.053	39178	897150		

File: 99022521.D21 Sample: BLANK PREP

1714

```

=====
Sample Name: S98T002849 SAM                      Date: 02/25/1999 15:30:52
Data File  : C:\DX\DATA\99022521.D22
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 22              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----

```

```

External      1      21      5700      5Hz      0.00      19.00      16000

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	7549	307212	1	
2	1.67		0.000	2900	22244	1	
3	2.20		0.000	4963	99884	1	
4	2.81		0.000	1243	26037	2	
8	3.40		0.000	1760	19702	2	
10	3.81		0.000	2381	17794	2	
11	3.95		0.000	2408	18747	2	
12	4.11		0.000	2539	24649	2	
14	4.40		0.000	2217	32222	2	
15	4.87		0.000	3849	56496	2	
16	4.99		0.000	3713	43618	2	
17	5.29		0.000	3293	31896	2	
18	5.44		0.000	3842	17830	2	
19	5.57		0.000	3929	53551	2	
20	5.77		0.000	3608	17393	2	
21	5.89		0.000	3712	58286	2	
22	7.12	fluoride	9.793	1476235	21163614	1	0.00
23	8.27	acetate	4.515	87385	1802597	2	0.00
24	8.93	formate	2.632	68301	1323085	2	0.00
25	9.71		0.000	7443	132126	1	

```

-----
Totals      16.940      1693271      25268982

```

File: 99022521.D22 Sample: S98T002849 SAM

1716


```

=====
Sample Name: S98T002849  DUP                      Date: 02/25/1999 15:54:04
Data File  : C:\DX\DATA\99022521.D23
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 23                Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700    5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	0.37		0.000	3737	46807	3	
4	1.17		0.000	1462	19588	1	
5	1.67		0.000	3960	44972	1	
6	2.21		0.000	5300	89196	1	
7	2.71		0.000	2846	44087	2	
13	3.91		0.000	1736	17666	2	
21	5.09		0.000	2721	35244	2	
22	5.39		0.000	2611	43201	2	
23	5.80		0.000	3552	60374	2	
24	7.12	fluoride	4.574	623463	9255513	2	0.00
25	8.24	acetate	3.971	73271	1556717	2	0.00
26	8.93	formate	2.179	51694	980830	2	0.00
27	9.73		0.000	10162	177100	1	
Totals			10.724	786514	12371294		

File: 99022521.D23 Sample: S98T002849 DUP

1717

```

=====
Sample Name: S98T002932 SAM                      Date: 02/25/1999 16:14:13
Data File  : C:\DX\DATA\99022521.D24
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 24              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          21  5700  5Hz  0.00 19.00    16000
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	2532	54400	1	
2	1.70		0.000	17231	184308	1	
3	2.63		0.000	4495	27932	2	
4	2.73		0.000	5422	38343	2	
5	2.88		0.000	5617	43802	2	
6	3.00		0.000	5481	68539	2	
7	3.21		0.000	4596	30825	2	
8	3.39		0.000	3703	31985	2	
9	3.52		0.000	3682	24110	2	
10	3.65		0.000	2361	16893	2	
11	3.81		0.000	2492	27298	2	
12	4.05		0.000	1879	25262	2	
18	5.47		0.000	1745	23260	2	
19	5.67		0.000	2177	22261	2	
20	5.88		0.000	2861	31747	2	
22	7.09	fluoride	10.404	1490439	22565419	2	0.00
23	8.27	acetate	1.492	19913	435278	2	0.00
24	8.96	formate	1.315	17058	328209	2	0.00
25	9.73		0.000	4947	107517	2	
Totals			13.210	1598631	24087388		

File: 99022521.D24 Sample: S98T002932 SAM

```

=====
Sample Name: S98T002932 DUP                      Date: 02/25/1999 16:34:28
Data File  : C:\DX\DATA\99022521.D25
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 25              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          21    5700  5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	1877	45268	1	
2	1.70		0.000	13578	128392	1	
3	2.10		0.000	1614	18348	1	
4	2.75		0.000	5767	92536	2	
5	3.03		0.000	5071	102630	2	
6	3.37		0.000	2324	27133	2	
20	5.61		0.000	1606	26043	2	
21	5.91		0.000	2223	32964	2	
24	7.09	fluoride	10.242	1462475	22195339	2	0.00
25	8.27	acetate	1.374	17641	382049	2	0.00
26	8.96	formate	1.286	16324	306579	2	0.00
27	9.65		0.000	5328	111805	2	
Totals			12.903	1535829	23469086		

File: 99022521.D25 Sample: S98T002932 DUP

1719

```

=====
Sample Name: S98T002932 SPK                      Date: 02/25/1999 16:54:06
Data File  : C:\DX\DATA\99022521.D26
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 26              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          21    5700  5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	4697	31278	2	
3	1.73		0.000	17659	165000	1	
4	2.96		0.000	5813	199494	2	
5	3.39		0.000	2092	16993	2	
12	4.92		0.000	1374	17363	2	
18	7.12	fluoride	24.127	3642161	54478989	2	0.00
19	7.81	glycolate	20.668	505441	9003674	2	0.00
20	8.27	acetate	23.461	531652	10314946	2	0.00
21	8.96	formate	25.095	978033	18530376	3	0.00
22	9.68		0.000	6067	100974	4	
Totals			93.350	5694990	92859087		

File: 99022521.D26 Sample: S98T002932 SPK

1720

LABCORE Completed Worklist Report for Worklist# 28347

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-O

Worklist Comment: Was originally 27982 BY105 rerun

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	8.40e-02	0.084	ug/mL
1 CCB	0		@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0		@IC4G-QC F*4	QC	6.10e1	6.31e+01	103.443 %	Recovery
2 LCS-INST	0		@IC4G-QC ACETATE2	QC	1.16e2	1.17e+02	100.862 %	Recovery
2 LCS-INST	0		@IC4G-QC FORMATE2	QC	1.19e2	1.27e+02	106.723 %	Recovery
2 LCS-INST	0		@IC4G-QC GLYCOLT1	QC	1.08e2	1.06e+02	98.148 %	Recovery
3 CCV	0		@IC4G-QC F*4	QC	6.29e1	5.93e+01	94.277 %	Recovery
3 CCV	0		@IC4G-QC ACETATE2	QC	1.20e2	1.21e+02	100.833 %	Recovery
3 CCV	0		@IC4G-QC FORMATE2	QC	1.20e2	1.29e+02	107.500 %	Recovery
3 CCV	0		@IC4G-QC GLYCOLT1	QC	1.03e2	1.00e+02	97.087 %	Recovery
4 BLNK-PREP	0		@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 ACETATE2	SOLID	1	3.20e-02	0.032	ug/g
4 BLNK-PREP	0		@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002714	0 W	@IC4G-01 F*4-01	SOLID	N/A	1.242e+04	127.700	ug/g
6 DUP	S98T002714	0 W	@IC4G-01 F*4-01	SOLID	1.24e+04	7.97e+03	43.495	RPD
7 SAMPLE	S98T002741	0 W	@IC4G-01 ACETATE2	SOLID	N/A	4.809e+02	84.130	ug/g
8 DUP	S98T002741	0 W	@IC4G-01 ACETATE2	SOLID	4.81e+02	4.69e+02	2.526	RPD
9 SPK	S98T002741	0 W	@IC4G-01 ACETATE2	SOLID	1.16e2	1.11e+02	95.690 %	Recovery

Final page for worklist# 28347

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 3/2/99

LABCORE Data Entry Template for Worklist# 28347 Analyst: ADP Instrument: IC 4552 Book# 59N21A LCSMethod: LA-533-105 Rev/Mod F-0 68N21 A CCV

Worklist Comment: Was originally 27982 BY105 rerun

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002714	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: F*4-01					
6 DUP	S98T002714	0 W	@IC4G-01	SOLID		
7 SAMPLE	S98T002741	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2					
8 DUP	S98T002741	0 W	@IC4G-01	SOLID		
9 SPK	S98T002741	0 W	@IC4G-01	SOLID		

Final page for worklist # 28347 See original
Signature _____ Date _____

Signature _____ Date _____

Data Entry Comments:

Uploaded 2/26/99 Validated 2/26/99 JM Luyt

LABCORE Data Entry Template for Worklist# 28347

Analyst: ADP Instrument: IC 4552 Book# 591021A LCS
Method: LA-533-105 Rev/Mod E-0 58N21A CCV

Worklist Comment: Was originally 27982 BY105 rerun

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002714	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: F*4-01					
6 DUP	S98T002714	0 W	@IC4G-01	SOLID		
7 SAMPLE	S98T002741	0 W	@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ACETATE2					
8 DUP	S98T002741	0 W	@IC4G-01	SOLID		
9 SPK	S98T002741	0 W	@IC4G-01	SOLID		
10 SAMPLE	S98T002848	0 W	@IC4G-01	SOLID	98000437	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01, GLYCOLT1					
11 DUP	S98T002848	0 W	@IC4G-01	SOLID		
12 SPK	S98T002848	0 W	@IC4G-01	SOLID		

Final page for worklist # 28347

See attached original
Signature _____ Date _____

Signature _____ Date _____

WKL

Data Entry Comments:

Replaced 2-24-99 -> cleared

Dr. Wawell

28347FEB.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

01/22/99 10:29
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27982

Analyst: ADP Instrument: IC 4552 Book# 59N21-A LCS
58N21-A CCV

Method: LA-533-105 Rev/Mod F-D

Worklist Comment: BY105 RERUN DF=21 JHW

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC-QC	QC		
2	LCS-INST				@IC-QC	QC		
3	CCV				@IC-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002714	0	W	@IC4G-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: F*4-01						
6	DUP	S98T002714	0	W	@IC4G-01	SOLID		
7	SAMPLE	S98T002741	0	W	@IC4G-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: ACETATE2						
8	DUP	S98T002741	0	W	@IC4G-01	SOLID		
9	SPK	S98T002741	0	W	@IC4G-01	SOLID		
10	SAMPLE	S98T002848	0	W	@IC4G-01	SOLID	98000437	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, GLYCOLT1						
11	DUP	S98T002848	0	W	@IC4G-01	SOLID		
12	SPK	S98T002848	0	W	@IC4G-01	SOLID		

} - Wrong Tray

Final page for worklist # 27982

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

uploaded 2-19-99

JH Wauell

27982 FEB. CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1724

Data Reprocessed On 02/18/1999 03:13:54

Sample Name: 59N21-A LCS Date: 02/18/1999 03:11:01
 Data File : C:\DX\DATA\99021711.D20
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 20 Detector: CDM-2
 Analyst : *Anthony Paruta* Column: AG14A-SC, AS14A-SC, SRS

02-18-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	1237	32060	1	
5	2.75		0.000	3312	46845	2	
6	2.85		0.000	3216	21026	2	
7	3.11		0.000	4879	76271	2	
8	3.28		0.000	4490	34776	2	
9	3.48		0.000	4472	38672	2	
10	3.60		0.000	4981	28154	2	
11	3.75		0.000	5441	160616	2	
12	4.23		0.000	5197	62691	2	
13	4.57		0.000	5538	116552	2	
14	4.84		0.000	4687	43104	2	
15	5.15		0.000	4612	95832	2	
16	5.48		0.000	3988	37988	2	
17	5.63		0.000	4488	124697	2	
18	6.69	fluoride	63.139103.51	2069734	28834822	2	0.00
19	7.33	glycolate	106.04078.19	564575	9630899	2	0.00
20	7.79	acetate	116.644100.56	580400	10668758	2	0.00
21	8.43	formate	127.448107.10	1067242	19622126	3	0.00
Totals			413.271	4342491	69675889		

File: 99021711.D20 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1725 TO 1734.

Data Reprocessed On 02/18/1999 03:33:54

```

=====
Sample Name: 58N21-A CCV                               Date: 02/18/1999 03:30:52
Data File  : C:\DX\DATA\99021711.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 21                    Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	10528	413060	1	
4	2.85		0.000	2233	41488	2	
5	3.07		0.000	2207	22797	2	
26	6.72	fluoride	59.327	1966657	27007595	2	0.00
27	7.36	glycolate	100.078	532777	9067282	2	0.00
28	7.81	acetate	120.925	610425	11066190	2	0.00
29	8.45	formate	129.170	1098545	19900703	2	0.00
Totals			409.500	4223372	67519115		

File: 99021711.D21 Sample: 58N21-A CCV

1726

```

=====
Sample Name: CCB                      Date: 02/18/1999 11:02:32
Data File  : C:\DX\DATA\99021801.D28
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 28   Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1  5700  5Hz   0.00 19.00   16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	32473	353332	1	
3	2.35		0.000	2929	20120	1	
11	3.53		0.000	2471	24735	2	
16	4.91		0.000	1863	27600	1	
21	6.25		0.000	1241	35670	3	
23	8.80		0.000	27908	2932087	2	
24	9.44	formate	0.084	16469	674869	2	0.00
25	9.95		0.000	3555	213109	2	
Totals			0.084	88909	4281521		

File: 99021801.D28 Sample: CCB

1727

```

=====
Sample Name: PREP BLANK                      Date: 02/18/1999 03:51:02
Data File  : C:\DX\DATA\99021711.D22
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 22
Analyst    :                               Detector: CDM-2
                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700  5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	9718	345455	2	
2	2.25		0.000	3780	56415	2	
3	2.55		0.000	4344	50145	2	
4	2.68		0.000	3132	32485	2	
5	2.87		0.000	1498	20969	2	
19	5.11		0.000	2375	31541	2	
20	5.53		0.000	1521	16109	2	
22	6.08		0.000	1783	28028	2	
24	6.75		0.000	1231	20656	1	
25	7.12		0.000	958	25802	2	
28	8.40	acetate	0.032	2554	62247	2	0.00
Totals			0.032	32894	689851		

File: 99021711.D22 Sample: PREP BLANK

1728

Data Reprocessed On 02/25/1999 16:46:57

```

=====
Sample Name: S98T002714 SAM                      Date: 02/18/1999 04:11:14
Data File  : E:\DATA\99021711.D23
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 23              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 5700 5Hz 5.77 9.70 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	9596	343612	1	
3	2.20		0.000	4099	63928	1	
4	2.79		0.000	6579	99391	2	
5	2.95		0.000	6488	63620	2	
6	3.07		0.000	5559	36813	2	
7	3.20		0.000	5336	70469	2	
8	3.48		0.000	4035	35581	2	
9	3.57		0.000	3535	17383	2	
10	3.65		0.000	3129	23209	2	
11	3.84		0.000	2509	21553	2	
14	4.31		0.000	1410	16411	2	
15	4.53		0.000	2284	19719	2	
16	4.71		0.000	2567	22903	2	
17	4.88		0.000	4230	49758	2	
18	5.04		0.000	4680	34179	2	
19	5.35		0.000	6248	191639	2	
20	6.72	fluoride	63.309	10324536	150682162	2	0.00
21	7.79	acetate	5.626	108663	2304115	2	0.00
22	8.43	formate	2.585	63190	1287456	2	0.00
23	9.17		0.000	21169	499420	2	
24	9.65		0.000	5112	79670	2	
Totals			71.520	10594953	155962990		

Data Reprocessed On 02/25/1999 16:40:54

```

=====
Sample Name: S98T002714  DUP                      Date: 02/18/1999 04:38:07
Data File  : E:\DATA\99021711.D24
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 24                Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700    5Hz    0.00  19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	3494	56681	2	
3	1.70		0.000	7273	66736	1	
4	2.13		0.000	3666	59514	1	
5	2.85		0.000	5582	113582	2	
6	3.00		0.000	5251	61689	2	
7	3.17		0.000	4864	40143	2	
8	3.33		0.000	3785	21409	2	
9	3.41		0.000	3815	48172	2	
10	3.72		0.000	1645	17735	2	
17	4.91		0.000	1276	23750	2	
18	5.40		0.000	3103	39659	2	
19	5.53		0.000	2776	18179	2	
20	6.69	fluoride	41.157	6707916	95308958	2	0.00
21	7.76	acetate	3.770	83904	1465765	2	0.00
22	8.43	glycolate	2.124	48849	752582	1	8.59
23	9.25		0.000	2374	105872	1	
Totals			47.051	6889572	98200427		

File: 99021711.D24 Sample: S98T002714 DUP

Data Reprocessed On 02/25/1999 16:45:59

```

=====
Sample Name: S98T002741 SAM                      Date: 02/18/1999 04:58:12
Data File  : E:\DATA\99021711.D25
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 25              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      21    5700    5Hz    5.53  10.90      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	7318	287149	2	
3	2.19		0.000	3992	65196	1	
4	2.80		0.000	6346	108422	2	
5	2.95		0.000	5326	69508	2	
6	3.21		0.000	3257	40078	3	
22	5.61		0.000	2700	28725	2	
24	6.69	fluoride	14.519	2271776	32048495	2	0.00
25	7.76	acetate	2.401	37891	846550	2	0.00
26	8.43	formate	2.055	40909	887127	3	0.00
27	9.17		0.000	6277	117786	4	
Totals			18.975	2385791	34499036		

File: 99021711.D25 Sample: S98T002741 SAM

Data Reprocessed On 02/25/1999 16:49:45

```

=====
Sample Name: S98T002741  DUP                      Date: 02/18/1999 05:18:25
Data File  : E:\DATA\99021711.D26
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 26
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz  5.37 10.66    16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	2815	45828	1	
3	1.63		0.000	3718	36146	1	
4	2.19		0.000	4610	72588	1	
5	2.69		0.000	6019	80478	2	
6	2.89		0.000	6193	62469	2	
7	2.96		0.000	5958	41505	2	
8	3.17		0.000	7778	122338	2	
9	3.60		0.000	3196	55714	2	
22	6.67	fluoride	13.748	2141927	30267182	2	0.00
23	7.73	acetate	2.322	39331	810890	2	0.00
24	8.40	formate	2.233	46628	1021342	3	0.00
25	9.15		0.000	5957	111258	4	
26	9.63		0.000	1700	16808	4	
Totals			18.303	2275830	32744546		

File: 99021711.D26 Sample: S98T002741 DUP

Data Reprocessed On 02/25/1999 16:54:47

```

=====
Sample Name: S98T002741  SPK                      Date: 02/18/1999 05:38:10
Data File  : E:\DATA\99021711.D27
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 27              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          21    5700  5Hz   5.21  9.86    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	9427	365113	2	
2	1.70		0.000	4144	32895	2	
3	2.19		0.000	3714	46340	1	
4	2.75		0.000	5224	81944	2	
5	3.08		0.000	6192	181215	2	
6	3.49		0.000	2397	16781	2	
17	5.48		0.000	2433	36957	2	
19	6.69	fluoride	28.669	4603835	65231849	2	0.00
20	7.33	glycolate	21.478	548627	9371525	2	0.00
21	7.76	acetate	24.605	586621	10825902	2	0.00
22	8.43	formate	26.518	1093750	19636512	3	0.00
23	9.15		0.000	7686	147312	4	
24	9.65		0.000	2195	28588	4	
Totals			101.269	6876244	106002934		

File: 99021711.D27 Sample: S98T002741 SPK

LABCORE Completed Worklist Report for Worklist# 28349

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: Was 27905 BY105 rerun, see original for signatures.jmf

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0	IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0	IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1	CCB	0	IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0	IC4G-QC F*4	QC	6.10e1	6.41e+01	105.082 % Recovery	
2	LCS-INST	0	IC4G-QC ACETATE2	QC	1.16e2	1.17e+02	100.862 % Recovery	
2	LCS-INST	0	IC4G-QC FORMATE2	QC	1.19e2	1.28e+02	107.563 % Recovery	
2	LCS-INST	0	IC4G-QC GLYCOLT1	QC	1.08e2	1.09e+02	100.926 % Recovery	
3	CCV	0	IC4G-QC F*4	QC	6.29e1	5.75e+01	91.415 % Recovery	
3	CCV	0	IC4G-QC ACETATE2	QC	1.20e2	1.19e+02	99.167 % Recovery	
3	CCV	0	IC4G-QC FORMATE2	QC	1.20e2	1.23e+02	102.500 % Recovery	
3	CCV	0	IC4G-QC GLYCOLT1	QC	1.03e2	9.53e+01	92.524 % Recovery	
4	SAMPLE	S98T002939	0	IC4G-01 F*4-01	LIQUID	N/A	6.308e+01	18.790 ug/mL
4	SAMPLE	S98T002939	0	IC4G-01 ACETATE2	LIQUID	N/A	2.367e+03	35.150 ug/mL
4	SAMPLE	S98T002939	0	IC4G-01 FORMATE2	LIQUID	N/A	1.352e+03	37.570 ug/mL
4	SAMPLE	S98T002939	0	IC4G-01 GLYCOLT1	LIQUID	N/A	< 3.394e+01	33.940 ug/mL
5	DUP	S98T002939	0	IC4G-01 F*4-01	LIQUID	6.31e+01	6.36e+01	0.789 RPD
5	DUP	S98T002939	0	IC4G-01 ACETATE2	LIQUID	2.37e+03	2.37e+03	0.000 RPD
5	DUP	S98T002939	0	IC4G-01 FORMATE2	LIQUID	1.35e+03	1.35e+03	0.000 RPD
5	DUP	S98T002939	0	IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD
6	SPK	S98T002939	0	IC4G-01 F*4-01	LIQUID	6.10e1	4.44e+01	72.787 % Recovery
6	SPK	S98T002939	0	IC4G-01 ACETATE2	LIQUID	1.16e2	1.20e+02	103.448 % Recovery
6	SPK	S98T002939	0	IC4G-01 FORMATE2	LIQUID	1.19e2	1.36e+02	114.206 % Recovery
6	SPK	S98T002939	0	IC4G-01 GLYCOLT1	LIQUID	1.08e2	1.16e+02	107.407 % Recovery
7	SAMPLE	S98T002964	0	IC4G-01 F*4-01	LIQUID	N/A	1.714e+02	18.790 ug/mL
7	SAMPLE	S98T002964	0	IC4G-01 ACETATE2	LIQUID	N/A	2.823e+03	35.150 ug/mL
7	SAMPLE	S98T002964	0	IC4G-01 FORMATE2	LIQUID	N/A	1.769e+03	37.570 ug/mL
7	SAMPLE	S98T002964	0	IC4G-01 GLYCOLT1	LIQUID	N/A	< 3.394e+01	33.940 ug/mL
8	DUP	S98T002964	0	IC4G-01 F*4-01	LIQUID	1.71e+02	2.15e+02	22.798 RPD
8	DUP	S98T002964	0	IC4G-01 ACETATE2	LIQUID	2.82e+03	2.75e+03	2.513 RPD
8	DUP	S98T002964	0	IC4G-01 FORMATE2	LIQUID	1.77e+03	1.72e+03	2.865 RPD
8	DUP	S98T002964	0	IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1	RPD

Final page for worklist# 28349

Analyst Signature

Date

Analyst Signature

Date

Jan 25 2:25/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

02/24/99 10:52

ws2

LABCORE Data Entry Template for Worklist# 28349

Analyst: ADP Instrument: IC 4552 Book# 59N21A LCS
 Method: LA-533-105 Rev/Mod F-0 58N21A CCV

Worklist Comment: Was 27905 BY105 rerun, see original for signatures.jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T002939 0		@IC4G-01 LIQUID		98000490 BY-105 (2)	
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
5 DUP	S98T002939 0		@IC4G-01 LIQUID			
6 SPK	S98T002939 0		@IC4G-01 LIQUID			
7 SAMPLE	S98T002964 0		@IC4G-01 LIQUID		98000509 BY-105 (2)	
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
8 DUP	S98T002964 0		@IC4G-01 LIQUID			

Final page for worklist # 28349

See original
 Signature _____ Date _____

Signature _____ Date _____

Data Entry Comments:

uploaded 2-24-99

JH Warrall

28349FEB.CSV

validated 2/25/99 JM Luge

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1736

LABCORE Data Entry Template for Worklist# 27905

Analyst: ADD Instrument: IC _____ Book# 59N21-A LES
Method: LA-533-105 Rev/Mod F-0 58N21-A CCV

Worklist Comment: BY105, @IC4G-01, Use 606 DF skm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4-QC	QC	} Wrong Tray	
2 ICV			@IC4-QC	QC		
3 CCV			@IC4-QC	QC		
4 SAMPLE	S98T002939 0		@IC4G-01	LIQUID	98000490	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
5 DUP	S98T002939 0		@IC4G-01	LIQUID		
6 SPK	S98T002939 0		@IC4G-01	LIQUID		
7 SAMPLE	S98T002964 0		@IC4G-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1						
8 DUP	S98T002964 0		@IC4G-01	LIQUID		

Final page for worklist # 27905

Anthony P. Smith 02-19-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

uploaded 2-19-99

JH Howell

27905 PER.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

=====

Sample Name: 59N21-A LCS	Date: 02/19/1999 01:03:56
Data File : C:\DX\DATA\99021811.D13	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System 2 Inject#: 13	Detector: CDM-2
Analyst : <i>Anthony P...</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

02-19-99

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	2340	81616	1	
31	7.36	fluoride	64.077105 ⁰⁴	1951930	29284614	2	0.00
32	8.08	glycolate	109.014100 ⁹⁴	533930	9912712	2	0.00
33	8.56	acetate	117.440101 ²⁴	543863	10742685	2	0.00
34	9.31	formate	127.669106 ⁷⁸	1020415	19657931	2	0.00
Totals			418.201	4052478	69679558		

File: 99021811.D13 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1738 TO 1745.

```

=====
Sample Name: 58N21-A CCV                      Date: 02/19/1999 02:03:08
Data File  : C:\DX\DATA\99021811.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 15           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 101 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	10846	625608	1	
5	3.04		0.000	2453	58765	2	
13	4.84		0.000	1440	29504	2	
15	5.44		0.000	1359	23056	2	
17	6.09		0.000	947	19866	2	
19	7.39	fluoride	57.496	1740160	26130709	2	0.00
20	8.11	glycolate	95.321	475157	8618715	2	0.00
21	8.59	acetate	119.017	544156	10889122	2	0.00
22	9.33	formate	122.629	996853	18842981	2	0.00
Totals			394.464	3773371	65238326		

File: 99021811.D15 Sample: 58N21-A CCV

```

=====
Sample Name: BLANK                      Date: 02/19/1999 00:43:29
Data File  : C:\DX\DATA\99021811.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12
Analyst    :                          Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	41257	555578	1	
2	2.19		0.000	15205	563917	2	
3	2.80		0.000	9110	213806	3	
5	3.75		0.000	663	38440	1	
8	5.24		0.000	14333	368340	1	
10	6.11		0.000	1923	24113	2	
11	6.41		0.000	4161	17679	2	
12	6.48		0.000	4787	181059	2	
14	8.99		0.000	2467	43380	1	
Totals			0.000	93906	2006312		

File: 99021811.D12 Sample: BLANK

```

=====
Sample Name: S98T002939 SAM                      Date: 02/19/1999 02:25:42
Data File  : C:\DX\DATA\99021811.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 606 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	4517	140776	2	
3	1.70		0.000	217660	2623080	1	
4	3.51		0.000	20386	1407808	2	
5	3.95		0.000	12714	72813	2	
6	4.08		0.000	11234	148061	2	
7	4.31		0.000	7050	45970	2	
8	4.43		0.000	5408	61825	3	
10	4.84		0.000	1677	58817	2	
11	5.24		0.000	10650	1016660	2	
12	7.23	fluoride	63.081	200153	3845595	2	0.00
13	8.40	acetate	2367.440	1507862	36051217	2	0.00
14	9.12	formate	1352.489	1764535	35682472	2	0.00
15	9.92		0.000	261735	4850871	1	
Totals			3783.009	4025581	86005967		

File: 99021811.D16 Sample: S98T002939 SAM

SS .100-10-2-10

1741

```

=====
Sample Name: S98T002939  DUP                               Date: 02/19/1999 02:45:41
Data File  : C:\DX\DATA\99021811.D17
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 17                      Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606    5700    5Hz    0.00 19.00    16000
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	8404	310666	1	
2	1.70		0.000	213736	2486086	1	
3	3.24		0.000	20040	687442	2	
4	3.37		0.000	18558	123496	2	
5	3.76		0.000	12907	438317	2	
6	4.01		0.000	8977	109023	2	
7	4.27		0.000	5167	80493	2	
10	4.61		0.000	1344	18338	2	
11	4.88		0.000	6653	41537	2	
12	4.99		0.000	8985	118538	2	
13	5.17		0.000	11805	173406	2	
14	5.48		0.000	17198	907189	2	
15	7.23	fluoride	63.603	205285	3886461	2	0.00
16	8.40	acetate	2365.221	1513428	36017990	2	0.00
17	9.12	formate	1348.273	1769717	35565756	2	0.00
18	9.92		0.000	264685	4932983	1	
Totals			3777.097	4086890	85897722		

File: 99021811.D17 Sample: S98T002939 DUP

SS .100-10-2-10

```

=====
Sample Name: S98T002939  SPK                      Date: 02/19/1999 03:05:29
Data File  : C:\DX\DATA\99021811.D18
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 18              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          606    5700  5Hz   0.00 19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	7243	283066	1	
2	1.70		0.000	214470	2529302	1	
3	3.24		0.000	21519	785551	2	
4	3.45		0.000	18320	224417	2	
5	3.60		0.000	16264	82314	2	
6	3.69		0.000	14823	177024	2	
7	3.92		0.000	10875	151035	2	
8	4.33		0.000	3305	74892	3	
11	4.92		0.000	5778	116476	2	
12	5.19		0.000	10577	75934	2	
13	5.41		0.000	14961	175318	2	
14	5.51		0.000	16285	108445	2	
15	5.63		0.000	18193	676277	2	
16	7.23	fluoride	284.874	1266512	21345450	2	0.00
17	7.92	glycolate	579.290	506798	8734344	2	0.00
18	8.40	acetate	2965.056	1915781	44949966	2	0.00
19	9.12	formate	2031.284	2689106	54783474	2	0.00
20	9.92		0.000	267145	5075821	2	
Totals			5860.503	7017954	140349105		

File: 99021811.D18 Sample: S98T002939 SPK

Data Reprocessed On 02/19/1999 10:40:10

```

=====
Sample Name: S98T002964 SAM                      Date: 02/19/1999 03:25:20
Data File  : E:\DATA\99021811.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          606   5700  5Hz   0.00 19.00      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	8378	283184	2	
2	1.70		0.000	237387	2704695	2	
3	3.19		0.000	20893	645146	2	
4	3.33		0.000	17804	175048	2	
5	3.49		0.000	14133	127228	2	
6	3.69		0.000	9493	146137	2	
7	3.91		0.000	4431	35128	3	
16	5.21		0.000	5602	123163	2	
17	5.47		0.000	11152	626275	2	
18	7.20	fluoride	171.422	709815	12363016	2	0.00
19	8.37	acetate	2822.607	1733288	42837929	2	0.00
20	9.09	formate	1769.438	2273890	47341481	2	0.00
21	9.89		0.000	360481	7010860	2	
Totals			4763.467	5406746	114419289		

File: 99021811.D19 Sample: S98T002964 SAM

Data Reprocessed On 02/19/1999 10:40:53

```

=====
Sample Name: S98T002964  DUP                      Date: 02/19/1999 04:02:34
Data File  : E:\DATA\99021811.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 20              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1          606   5700  5Hz   0.00  19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	4968	139982	1	
2	1.70		0.000	264350	2986738	1	
4	5.87		0.000	131976	27143286	2	
5	7.20	fluoride	215.085	763332	15812234	2	0.00
6	8.37	acetate	2752.678	1619702	41799057	2	0.00
7	9.07	formate	1722.115	2163011	46006564	2	0.00
8	9.87		0.000	377745	7736612	2	
Totals			4689.878	5325084	141624474		

File: 99021811.D20 Sample: S98T002964 DUP

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 28394

Analyst: krm

Instrument: IC40S2

Book#: 61N20A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 FOR @IC-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC-QC	F	QC	1	<1.20e-2		ug/mL
1 CCB	0	@IC-QC	CL	QC	1	2.20e-02	0.022	ug/mL
1 CCB	0	@IC-QC	NO2	QC	1	<1.08e-1		ug/mL
1 CCB	0	@IC-QC	BR	QC	1	<1.25e-1		ug/mL
1 CCB	0	@IC-QC	NO3	QC	1	<1.39e-1		ug/mL
1 CCB	0	@IC-QC	PO4	QC	1	<1.20e-1		ug/mL
1 CCB	0	@IC-QC	SO4	QC	1	<1.38e-1		ug/mL
1 CCB	0	@IC-QC	OXALATE2	QC	1	<1.05e-1		ug/mL
2 LCS-INST	0	@IC-QC	F	QC	5.92e1	6.17e+01	104.223 % Recovery	
2 LCS-INST	0	@IC-QC	CL	QC	7.96e1	7.89e+01	99.121 % Recovery	
2 LCS-INST	0	@IC-QC	NO2	QC	5.53e2	4.93e+02	89.150 % Recovery	
2 LCS-INST	0	@IC-QC	BR	QC	5.86e2	5.81e+02	99.147 % Recovery	
2 LCS-INST	0	@IC-QC	NO3	QC	5.92e2	6.41e+02	108.277 % Recovery	
2 LCS-INST	0	@IC-QC	PO4	QC	5.48e2	5.36e+02	97.810 % Recovery	
2 LCS-INST	0	@IC-QC	SO4	QC	6.38e2	6.45e+02	101.097 % Recovery	
2 LCS-INST	0	@IC-QC	OXALATE2	QC	5.33e2	5.51e+02	103.377 % Recovery	
3 CCV	0	@IC-QC	F	QC	6.40e1	6.20e+01	96.875 % Recovery	
3 CCV	0	@IC-QC	CL	QC	8.95e1	9.36e+01	104.581 % Recovery	
3 CCV	0	@IC-QC	NO2	QC	5.23e2	5.56e+02	106.310 % Recovery	
3 CCV	0	@IC-QC	BR	QC	6.30e2	6.22e+02	98.730 % Recovery	
3 CCV	0	@IC-QC	NO3	QC	7.00e2	6.98e+02	99.714 % Recovery	
3 CCV	0	@IC-QC	PO4	QC	6.35e2	6.28e+02	98.898 % Recovery	
3 CCV	0	@IC-QC	SO4	QC	7.00e2	7.07e+02	101.000 % Recovery	
3 CCV	0	@IC-QC	OXALATE2	QC	5.30e2	5.36e+02	101.132 % Recovery	
4 BLNK-PREP	0	@IC-01	F-02	SOLID	1	<1.20e-2		ug/g
4 BLNK-PREP	0	@IC-01	CL-02	SOLID	1	3.80e-02	0.038	ug/g
4 BLNK-PREP	0	@IC-01	NO2-02	SOLID	1	<1.08e-1		ug/g
4 BLNK-PREP	0	@IC-01	BR-02	SOLID	1	<1.25e-1		ug/g
4 BLNK-PREP	0	@IC-01	NO3-02	SOLID	1	<1.39e-1		ug/g
4 BLNK-PREP	0	@IC-01	PO4-02	SOLID	1	<1.20e-1		ug/g
4 BLNK-PREP	0	@IC-01	SO4-02	SOLID	1	<1.38e-1		ug/g
4 BLNK-PREP	0	@IC-01	OXALATE2	SOLID	1	<1.05e-1		ug/g
5 SAMPLE	S98T002933	0 W	@IC-01	F-02	SOLID	N/A	7.698e+02	121.360 ug/g
6 DUP	S98T002933	0 W	@IC-01	F-02	SOLID	7.70e+02	7.23e+02	6.296 RPD
6 DUP	S98T002933	0 W	@IC-01	CL-02	SOLID	?	2.49e+02	RPD
6 DUP	S98T002933	0 W	@IC-01	NO2-02	SOLID	?	4.02e+03	RPD
6 DUP	S98T002933	0 W	@IC-01	BR-02	SOLID	?	<1.27e3	RPD
6 DUP	S98T002933	0 W	@IC-01	NO3-02	SOLID	?	6.10e+05	RPD
6 DUP	S98T002933	0 W	@IC-01	PO4-02	SOLID	?	<1.22e3	RPD
6 DUP	S98T002933	0 W	@IC-01	SO4-02	SOLID	?	4.35e+03	RPD
6 DUP	S98T002933	0 W	@IC-01	OXALATE2	SOLID	?	1.80e+03	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 28394

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 SPK	S98T002933	0 W	@IC-01 F-02	SOLID	5.90e01	6.25e+01	105.932 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 CL-02	SOLID	7.90e01	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 NO2-02	SOLID	5.44e02	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 BR-02	SOLID	5.89e02	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 NO3-02	SOLID	5.94e02	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 PO4-02	SOLID	5.45e02	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 SO4-02	SOLID	6.31e02	0.00e+00	0.000 %	Recovery
7 SPK	S98T002933	0 W	@IC-01 OXALATE2	SOLID	5.32e02	0.00e+00	0.000 %	Recovery
8 SAMPLE	S98T002934	0 W	@IC-01 F-02	SOLID	N/A	8.351e+02	119.600	ug/g
9 DUP	S98T002934	0 W	@IC-01 F-02	SOLID	8.35e+02	6.98e+02	17.873	RPD
9 DUP	S98T002934	0 W	@IC-01 CL-02	SOLID	?	9.48e+02		RPD
9 DUP	S98T002934	0 W	@IC-01 NO2-02	SOLID	?	1.14e+04		RPD
9 DUP	S98T002934	0 W	@IC-01 BR-02	SOLID	?	<1.25e3		RPD
9 DUP	S98T002934	0 W	@IC-01 NO3-02	SOLID	?	5.32e+05		RPD
9 DUP	S98T002934	0 W	@IC-01 PO4-02	SOLID	?	2.48e+03		RPD
9 DUP	S98T002934	0 W	@IC-01 SO4-02	SOLID	?	5.04e+03		RPD
9 DUP	S98T002934	0 W	@IC-01 OXALATE2	SOLID	?	1.77e+03		RPD
10 SPK	S98T002934	0 W	@IC-01 F-02	SOLID	5.92e1	6.28e+01	106.081 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 CL-02	SOLID	7.96e1	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 NO2-02	SOLID	5.53e2	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 BR-02	SOLID	5.86e2	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 NO3-02	SOLID	5.92e2	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 PO4-02	SOLID	5.48e2	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 SO4-02	SOLID	6.38e2	0.00e+00	0.000 %	Recovery
10 SPK	S98T002934	0 W	@IC-01 OXALATE2	SOLID	5.33e2	0.00e+00	0.000 %	Recovery
11 SAMPLE	S98T002935	0 W	@IC-01 F-02	SOLID	N/A	7.462e+02	123.400	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 CL-02	SOLID	N/A	8.748e+02	174.800	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 NO2-02	SOLID	N/A	1.121e+04	1110.000	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 BR-02	SOLID	N/A	1.285e+03	1285.000	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 NO3-02	SOLID	N/A	4.074e+05	1429.000	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 PO4-02	SOLID	N/A	<1.234e+03	1234.000	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 SO4-02	SOLID	N/A	6.750e+03	1419.000	ug/g
11 SAMPLE	S98T002935	0 W	@IC-01 OXALATE2	SOLID	N/A	7.862e+03	1080.000	ug/g
12 DUP	S98T002935	0 W	@IC-01 F-02	SOLID	7.46e+02	1.04e+03	32.923	RPD
12 DUP	S98T002935	0 W	@IC-01 CL-02	SOLID	8.75e+02	1.33e+03	41.270	RPD
12 DUP	S98T002935	0 W	@IC-01 NO2-02	SOLID	1.12e+04	1.39e+04	21.514	RPD
12 DUP	S98T002935	0 W	@IC-01 BR-02	SOLID	<1.28e3	<1.28e3		RPD
12 DUP	S98T002935	0 W	@IC-01 NO3-02	SOLID	4.07e+05	3.04e+05	28.973	RPD
12 DUP	S98T002935	0 W	@IC-01 PO4-02	SOLID	<1.23e3	2.82e+03		RPD
12 DUP	S98T002935	0 W	@IC-01 SO4-02	SOLID	6.75e+03	9.20e+03	30.721	RPD
12 DUP	S98T002935	0 W	@IC-01 OXALATE2	SOLID	7.86e+03	1.20e+04	41.692	RPD
13 SPK	S98T002935	0 W	@IC-01 F-02	SOLID	5.92e1	6.07e+01	102.534 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 CL-02	SOLID	7.96e1	7.67e+01	96.357 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 NO2-02	SOLID	5.53e2	4.89e+02	88.427 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 BR-02	SOLID	5.86e2	5.92e+02	101.024 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 NO3-02	SOLID	5.92e2	8.82e+02	148.986 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 PO4-02	SOLID	5.48e2	5.32e+02	97.080 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 SO4-02	SOLID	6.38e2	6.45e+02	101.097 %	Recovery
13 SPK	S98T002935	0 W	@IC-01 OXALATE2	SOLID	5.33e2	5.49e+02	103.002 %	Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28394

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 28394

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Jan Lye 3/12/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28394

Analyst: KRM Instrument: IC 4052 Book# GINDID 60N213

Method: LA-533-105 Rev/Mod _____

Worklist Comment: BY105 FOR @IC-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002933 0 W		@IC-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: F-02					
6 DUP	S98T002933 0 W		@IC-01	SOLID		
7 SPK	S98T002933 0 W		@IC-01	SOLID		
8 SAMPLE	S98T002934 0 W		@IC-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: F-02					
9 DUP	S98T002934 0 W		@IC-01	SOLID		
10 SPK	S98T002934 0 W		@IC-01	SOLID		
11 LAM. LE	S98T002935 0 W		@IC-01	SOLID	98000490	BY-105 (2)
	Analyte Requested: BR-02 , CL-02 , F-02 , NO2-02 , NO3-02 , OXALATE2, PO4-02 , SO4-02					
12 DUP	S98T002935 0 W		@IC-01	SOLID		
13 SPK	S98T002935 0 W		@IC-01	SOLID		

Final page for worklist # 28394

Signature _____ Date 2-27-99

Signature _____ Date _____

Data Entry Comments: Spiked with GINDIDuploaded 3-3-99
John Wainwright 28394FEB.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

RPD's still high
but probably due to
sample inhomogeneity
3/9/99 M. Luge

=====

Sample Name: CCB	Date: 02/27/1999 19:41:19
Data File : C:\DX\DATA\99022721.D01	
Method : C:\DX\METHOD\4000SYS2.MET	
ACI Address: 1 System: 2 Inject#: 1	Detector: CDM-1
Analyst : <i>[Signature]</i>	Column: AG4A/AS4A anion column

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 3000 5Hz 0.00 10.00 30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.31		0.000	16	70	2	
2	1.44	chloride	0.022	25	83	2	-0.69
Totals			0.022	41	154		

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1750 TO 1764

```

=====
Sample Name: LCS                      Date: 02/27/1999 19:55:45
Data File : C:\DX\DATA\99022721.D02
Method    : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 2      Detector: CDM-1
Analyst   :                          Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate Start Stop Area Reject
-----
External           1           101 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.42		0.000	1506	3052	1	
2	0.80		0.000	57	209	2	
3	0.95	fluoride	61.672	1799	5525	2	-1.39
4	1.43	chloride	78.886	1444	4657	1	-1.15
5	1.71	nitrite	492.872	4776	18553	1	-0.96
6	2.55	bromide	581.153	3177	15806	1	-0.26
7	2.90	nitrate	640.888	4003	23285	1	0.00
8	4.35	phosphate	536.206	757	8685	1	-0.53
9	5.65	sulfate	645.090	2350	28496	1	-0.99
10	7.36	oxalate	550.849	1192	19156	1	-1.47
Totals			3587.616	21062	127424		

File: 99022721.D02 Sample: LCS

Sample Name: CCV 60N21B Date: 02/27/1999 20:08:56
 Data File : C:\DX\DATA\99022721.D03
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 3 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	73	295	2	
2	0.95	fluoride	62.016	1814	5557	2	-1.39
3	1.43	chloride	93.626	1640	5549	1	-1.15
4	1.71	nitrite	555.717	5348	21017	1	-0.96
5	2.55	bromide	621.783	3429	16958	1	-0.52
6	2.89	nitrate	697.969	4275	25454	1	-0.23
7	4.35	phosphate	628.095	895	10228	1	-0.53
8	5.65	sulfate	706.913	2624	31325	1	-0.99
9	7.36	oxalate	536.350	1146	18642	1	-1.47
Totals			3902.468	21243	135024		

File: 99022721.D03 Sample: CCV 60N21B

Sample Name: PREP BLK Date: 02/27/1999 20:30:05
 Data File : C:\DX\DATA\99022721.D04
 Method : C:\DX\METHOD\4000SYS2.MET
 ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-1
 Analyst : Column: AG4A/AS4A anion column

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	3000	5Hz	0.00	10.00		30

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.23		0.000	33	306	1	
2	1.43	chloride	0.038	43	175	1	-1.15
Totals			0.038	76	481		

File: 99022721.D04 Sample: PREP BLK

```

=====
Sample Name: S98T002933      Date: 02/27/1999 21:37:44
Data File  : C:\DX\DATA\99022721.D08
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 8      Detector: CDM-1
Analyst    :                  Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate  Start  Stop Area Reject
-----
External           1           51  3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	111	469	2	
2	0.95	fluoride	3.885	219	595	2	-1.39
3	1.43	chloride	1.324	34	105	1	-1.61
4	1.71	nitrite	19.943	259	981	1	-1.35
5	2.79	nitrate	3092.065	41007	284902	1	-3.68
6	5.65	sulfate	25.491	128	1638	1	-0.99
7	7.36	oxalate	10.053	25	402	1	-1.47
Totals			3152.761	41783	289093		

File: 99022721.D08 Sample: S98T002933

```

=====
Sample Name: S98T002933 DUP                      Date: 02/27/1999 22:04:37
Data File  : C:\DX\DATA\99022721.D10
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 10              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate Start Stop Area Reject
-----
External           1           51  3000  5Hz  0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.79		0.000	105	424	2	
2	0.95	fluoride	3.640	191	551	2	-1.39
3	1.43	chloride	1.253	33	97	1	-1.61
4	1.71	nitrite	20.246	265	1004	1	-1.35
5	2.79	nitrate	3068.725	40392	281941	1	-3.91
6	5.65	sulfate	21.896	111	1320	1	-0.99
7	7.41	oxalate	9.054	24	332	1	-0.76
Totals			3124.815	41121	285670		

File: 99022721.D10 Sample: S98T002933 DUP

```

=====
Sample Name: S98T002933 SPK                      Date: 02/27/1999 21:53:39
Data File  : F:\DATA\99022721.D09
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 9                Detector:CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      51      3000  5Hz  0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	166	644	2	
2	0.95	fluoride	35.123	2025	6251	2	-1.39
3	1.43	chloride	40.280	1387	4710	1	-1.61
4	1.71	nitrite	260.793	5002	19477	1	-1.35
5	2.54	bromide	285.701	3348	15372	1	-0.78
6	2.78	nitrate	3324.189	44603	315732	1	0.00
7	4.35	phosphate	275.164	775	8832	1	-0.53
8	5.65	sulfate	343.231	2544	30080	1	-0.99
9	7.36	oxalate	293.567	1221	20238	1	-1.47
Totals			4858.047	61072	421335		

File: 99022721.D09 Sample: S98T002933 SPK

```

=====
Sample Name: S98T002934                      Date: 02/27/1999 22:22:14
Data File  : C:\DX\DATA\99022721.D11
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 11         Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate Start Stop Area Reject
-----
External           1           51 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	107	427	2	
2	0.95	fluoride	4.274	228	665	2	-1.39
3	1.43	chloride	4.727	157	503	1	-1.15
4	1.71	nitrite	62.954	1111	4255	1	-1.35
5	2.80	nitrate	2773.485	36372	246305	1	-3.45
6	4.35	phosphate	17.194	35	384	1	-0.53
7	5.65	sulfate	22.806	113	1400	1	-0.99
8	7.36	oxalate	11.583	25	509	1	-1.47
Totals			2897.023	38148	254449		

File: 99022721.D11 Sample: S98T002934

```

=====
Sample Name: S98T002934 DUP      Date: 02/27/1999 22:33:08
Data File  : C:\DX\DATA\99022721.D12
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 12      Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume  Dilution Points Rate  Start  Stop Area Reject
-----
External           1           51  3000  5Hz   0.00  10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	107	454	2	
2	0.95	fluoride	3.551	193	535	2	-1.39
3	1.43	chloride	4.820	160	514	1	-1.15
4	1.71	nitrite	57.839	977	3865	1	-1.35
5	2.81	nitrate	2705.227	35319	238494	1	-3.22
6	4.35	phosphate	12.613	24	236	1	-0.53
7	5.65	sulfate	25.649	112	1652	1	-0.99
8	7.41	oxalate	9.013	27	330	1	-0.76
Totals			2818.711	36919	246080		

File: 99022721.D12 Sample: S98T002934 DUP

```

=====
Sample Name: S98T002934 SPK                      Date: 02/27/1999 22:44:16
Data File  : C:\DX\DATA\99022721.D13
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 13              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51 3000 5Hz 0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	155	611	2	
2	0.95	fluoride	35.667	1959	6350	2	-1.39
3	1.43	chloride	42.998	1483	5036	1	-1.61
4	1.71	nitrite	302.909	5836	22754	1	-1.35
5	2.54	bromide	285.151	3289	15341	1	-0.78
6	2.79	nitrate	2992.951	39417	272483	1	-3.68
7	4.35	phosphate	277.968	794	8925	1	-0.53
8	5.60	sulfate	339.209	1969	29715	1	-1.93
9	7.31	oxalate	281.210	1069	19371	1	-2.19
Totals			4558.065	55972	380586		

File: 99022721.D13 Sample: S98T002934 SPK

```

=====
Sample Name: S98T002935                      Date: 02/27/1999 22:55:13
Data File  : C:\DX\DATA\99022721.D14
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 14          Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          51  3000  5Hz   0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.81		0.000	91	411	2	
2	0.95	fluoride	3.701	196	562	2	-1.39
3	1.43	chloride	4.339	139	458	1	-1.15
4	1.71	nitrite	55.611	932	3695	1	-1.35
5	2.82	nitrate	2020.878	25584	166960	1	-2.76
6	5.65	sulfate	33.479	192	2344	1	-0.99
7	7.36	oxalate	38.995	139	2420	1	-1.47
Totals			2157.004	27272	176850		

File: 99022721.D14 Sample: S98T002935

```

=====
Sample Name: S98T002935 DUP                      Date: 02/27/1999 23:10:03
Data File  : C:\DX\DATA\99022721.D15
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 15              Detector: CDM-1
Analyst    :                      Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51    3000  5Hz   0.00 10.00      30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	83	324	1	
2	0.95	fluoride	5.161	281	825	1	-1.39
3	1.43	chloride	6.633	186	726	1	-1.15
4	1.71	nitrite	69.336	1243	4741	1	-1.35
5	2.84	nitrate	1511.722	19264	119839	1	-2.07
6	4.35	phosphate	14.024	29	282	1	-0.53
7	5.65	sulfate	45.776	287	3433	1	-0.99
8	7.36	oxalate	59.615	245	3858	1	-1.47
Totals			1712.268	21618	134028		

File: 99022721.D15 Sample: S98T002935 DUP

```

=====
Sample Name: S98T002935 SPK                      Date: 02/27/1999 23:21:26
Data File  : C:\DX\DATA\99022721.D16
Method     : C:\DX\METHOD\4000SYS2.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-1
Analyst    :                               Column: AG4A/AS4A anion column
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          51  3000  5Hz  0.00 10.00          30
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.80		0.000	154	615	2	
2	0.95	fluoride	34.033	1960	6052	2	-1.39
3	1.43	chloride	42.707	1532	5001	1	-1.15
4	1.71	nitrite	299.882	5835	22518	1	-0.96
5	2.54	bromide	296.061	3276	15953	1	-0.78
6	2.81	nitrate	2462.115	31744	211775	1	-3.22
7	4.35	phosphate	266.241	772	8536	1	-0.53
8	5.60	sulfate	356.030	2109	31241	1	-1.93
9	7.31	oxalate	313.290	1204	21624	1	-2.19
Totals			4070.360	48586	323314		

File: 99022721.D16 Sample: S98T002935 SPK

LABCORE Completed Worklist Report for Worklist# 28395

Analyst: krm

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-O

Worklist Comment: BY105 FOR @IC4G-01 RLC

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	3.00e-02	0.030	ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	6.36e+01	104.262	% Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.13e+02	97.414	% Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.26e+02	105.882	% Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLTI	QC	1.08e2	1.08e+02	100.000	% Recovery
3 CCV	0	@IC4G-QC F*4	QC	6.23e1	5.92e+01	94.118	% Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.20e+02	100.000	% Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.26e+02	105.000	% Recovery
3 CCV	0	@IC4G-QC GLYCOLTI	QC	1.03e2	9.94e+01	96.505	% Recovery
4 SAMPLE	S98T002960	@IC4G-01 F*4-01	LIQUID	N/A	3.248e+02	34.440	ug/mL
5 DUP	S98T002960	@IC4G-01 F*4-01	LIQUID	3.25e+02	3.28e+02	0.919	RPD
5 DUP	S98T002960	@IC4G-01 ACETATE2	LIQUID	?	1.66e+03		RPD
5 DUP	S98T002960	@IC4G-01 FORMATE2	LIQUID	?	1.20e+03		RPD
5 DUP	S98T002960	@IC4G-01 GLYCOLTI	LIQUID	?	<6.22e1		RPD
6 SPK	S98T002960	@IC4G-01 F*4-01	LIQUID	6.10e1	6.65e+01	109.016	% Recovery
6 SPK	S98T002960	@IC4G-01 ACETATE2	LIQUID	1.16e2	0.00e+00	0.000	% Recovery
6 SPK	S98T002960	@IC4G-01 FORMATE2	LIQUID	1.19e2	0.00e+00	0.000	% Recovery
6 SPK	S98T002960	@IC4G-01 GLYCOLTI	LIQUID	1.08e2	0.00e+00	0.000	% Recovery
7 SAMPLE	S98T002961	@IC4G-01 F*4-01	LIQUID	N/A	3.169e+02	34.440	ug/mL
8 DUP	S98T002961	@IC4G-01 F*4-01	LIQUID	3.17e+02	3.24e+02	2.184	RPD
8 DUP	S98T002961	@IC4G-01 ACETATE2	LIQUID	?	1.97e+03		RPD
8 DUP	S98T002961	@IC4G-01 FORMATE2	LIQUID	?	1.39e+03		RPD
8 DUP	S98T002961	@IC4G-01 GLYCOLTI	LIQUID	?	<6.22e1		RPD
9 SPK	S98T002961	@IC4G-01 F*4-01	LIQUID	6.10e1	6.49e+01	106.393	% Recovery
9 SPK	S98T002961	@IC4G-01 ACETATE2	LIQUID	1.16e2	0.00e+00	0.000	% Recovery
9 SPK	S98T002961	@IC4G-01 FORMATE2	LIQUID	1.19e2	0.00e+00	0.000	% Recovery
9 SPK	S98T002961	@IC4G-01 GLYCOLTI	LIQUID	1.08e2	0.00e+00	0.000	% Recovery
10 SAMPLE	S98T002962	@IC4G-01 F*4-01	LIQUID	N/A	2.332e+02	34.440	ug/mL
10 SAMPLE	S98T002962	@IC4G-01 ACETATE2	LIQUID	N/A	2.064e+03	64.440	ug/mL
10 SAMPLE	S98T002962	@IC4G-01 FORMATE2	LIQUID	N/A	1.346e+03	68.880	ug/mL
10 SAMPLE	S98T002962	@IC4G-01 GLYCOLTI	LIQUID	N/A	< 6.22e+01	62.220	ug/mL
11 DUP	S98T002962	@IC4G-01 F*4-01	LIQUID	2.33e+02	2.37e+02	1.702	RPD
11 DUP	S98T002962	@IC4G-01 ACETATE2	LIQUID	2.06e+03	2.12e+03	2.871	RPD
11 DUP	S98T002962	@IC4G-01 FORMATE2	LIQUID	1.35e+03	1.39e+03	2.920	RPD
11 DUP	S98T002962	@IC4G-01 GLYCOLTI	LIQUID	<6.22e1	<6.22e1		RPD
12 SPK	S98T002962	@IC4G-01 F*4-01	LIQUID	6.10e1	6.24e+01	102.295	% Recovery
12 SPK	S98T002962	@IC4G-01 ACETATE2	LIQUID	1.16e2	1.21e+02	104.310	% Recovery
12 SPK	S98T002962	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.48e+02	124.370	% Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28395

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
12	SPK	S98T002962	0		@IC4G-01	GLYCOLT1 LIQUID	1.08e2	1.21e+02	112.037	% Recovery

Final page for worklist# 28395

_____
Analyst Signature   Date

_____
Analyst Signature   Date

3/12/99
Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 28395

Analyst: KRM Instrument: IC 4552 Book# 59N21A

Method: LA-533-105 Rev/Mod _____

Worklist Comment: BY105 FOR @IC4G-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 SAMPLE	S98T002960 0		@IC4G-01 LIQUID		98000509	BY-105 (2)
	Analytes Requested: F*4-01					
5 DUP	S98T002960 0		@IC4G-01 LIQUID			
6 SPK	S98T002960 0		@IC4G-01 LIQUID			
7 SAMPLE	S98T002961 0		@IC4G-01 LIQUID		98000509	BY-105 (2)
	Analytes Requested: F*4-01					
8 DUP	S98T002961 0		@IC4G-01 LIQUID			
9 SPK	S98T002961 0		@IC4G-01 LIQUID			
10 SAMPLE	S98T002962 0		@IC4G-01 LIQUID		98000509	BY-105 (2)
	Analytes Requested: ACETATE2, F*4-01 , FORMATE2, GLYCOLT1					
11 DUP	S98T002962 0		@IC4G-01 LIQUID			
12 SPK	S98T002962 0		@IC4G-01 LIQUID			

Final page for worklist # 28395

Signature _____ Date 2-27-99

Signature _____ Date _____

try comment: SPICED WITH 59N21Aupbeafal 373-99
Al Howell 28395FEB.CSVValidated 3/10/99 JMS

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Data Reprocessed On 03/03/1999 13:35:13

Sample Name: CCB Date: 02/27/1999 19:45:19
 Data File : E:\DATA\99022701.D18
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 18 Detector: CDM-2
 Analyst: *SNF/j2KRM* Column: AG14A-SC, AS14A-SC, SRS
3/25/99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	42991	1075366	2	
2	1.47		0.000	3933	69671	2	
13	7.28	fluoride	0.030	8685	350160	1	0.00
Totals			0.030	55609	1495197		

File: 99022701.D18 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1766 TO 1779

Data Reprocessed On 03/03/1999 13:34:54

```

=====
Sample Name: LCS 59N21A                      Date: 02/27/1999 20:08:07
Data File  : E:\DATA\99022701.D19
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 19          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External      1      101    5700    5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	6896	245176	1	
4	2.69		0.000	2281	26514	2	
6	3.04		0.000	2389	32863	2	
10	3.65		0.000	859	18100	1	
25	7.25	fluoride	63.597	1967609	29054497	2	0.00
26	7.95	glycolate	107.627	533346	9781247	2	0.00
27	8.43	acetate	113.448	541203	10371990	2	0.00
28	9.15	formate	126.113	1031888	19406183	2	0.00
Totals			410.786	4086470	68936571		

File: 99022701.D19 Sample: LCS 59N21A

```

=====
Sample Name: CCV 58N21A                      Date: 02/27/1999 20:32:42
Data File  : C:\DX\DATA\99022701.D20
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 20          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz   0.00 19.00   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	17391	826363	2	
4	2.73		0.000	2630	27853	2	
5	2.95		0.000	4122	42186	2	
6	3.11		0.000	3211	45017	2	
7	3.39		0.000	2336	30070	2	
17	5.41		0.000	1821	18931	2	
24	7.28	fluoride	59.168	1847502	26931414	2	0.00
25	7.97	glycolate	99.413	498905	9004479	2	0.00
26	8.43	acetate	120.115	556996	10991061	2	0.00
27	9.15	formate	126.211	1038934	19421972	2	0.00
Totals			404.907	3973847	67339344		

File: 99022701.D20 Sample: CCV 58N21A

```

=====
Sample Name: S98T002960                      Date: 02/27/1999 20:53:44
Data File  : C:\DX\DATA\99022701.D21
Method     : C:\DX\METHOD\MAS14GLY.MET
ACT Address: 1 System: 2 Inject#: 21          Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1111  5700  5Hz  0.00 19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	1974	32198	2	
2	0.83		0.000	1190	22855	2	
4	1.70		0.000	96620	1144444	1	
5	3.05		0.000	10542	335342	2	
6	3.28		0.000	8681	137046	2	
7	3.49		0.000	5917	91778	3	
10	4.24		0.000	2496	26687	2	
11	4.41		0.000	3376	21719	2	
12	4.63		0.000	5186	49725	2	
13	4.72		0.000	5747	124259	2	
14	5.11		0.000	7557	139272	2	
15	5.37		0.000	9339	124741	2	
16	5.68		0.000	10330	337509	2	
17	7.15	fluoride	324.789	803125	12815577	2	0.00
18	8.32	acetate	1632.594	631270	13613396	2	0.00
19	9.01	formate	1185.685	858628	16449723	2	0.00
20	9.81		0.000	137072	2591175	2	
Totals			3143.068	2599050	48057449		

File: 99022701.D21 Sample: S98T002960

```

=====
Sample Name: S98T002960 DUP                      Date: 02/27/1999 21:15:41
Data File  : C:\DX\DATA\99022701.D22
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 22              Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1111  5700  5Hz  0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	1772	31492	2	
4	1.70		0.000	101387	1217060	1	
5	2.93		0.000	12623	268781	2	
6	3.08		0.000	11317	131093	2	
7	3.25		0.000	8976	46114	2	
8	3.32		0.000	7991	56326	2	
9	3.49		0.000	6331	43675	2	
10	3.59		0.000	5072	29157	2	
11	3.73		0.000	2885	23791	2	
14	4.17		0.000	1709	16583	2	
15	4.36		0.000	3271	32813	2	
16	4.55		0.000	4092	28318	2	
17	4.68		0.000	4839	33270	2	
18	4.77		0.000	5410	78709	2	
19	5.03		0.000	6046	76858	2	
20	5.32		0.000	8037	106000	2	
21	5.53		0.000	9456	381379	2	
22	7.15	fluoride	327.634	810762	12938063	2	0.00
23	8.32	acetate	1661.973	640239	13860390	2	0.00
24	9.01	formate	1202.060	880785	16689377	2	0.00
25	9.79		0.000	141686	2761634	2	
Totals			3191.667	2674687	48850883		

```

=====
Sample Name: S98T002960 SPK                      Date: 02/27/1999 21:35:46
Data File  : C:\DX\DATA\99022701.D23
Method     : C:\DX\METHOD\MAS14GLV.MET
ACI Address: 1 System: 2 Inject#: 23              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1111  5700 5Hz  0.00 19.00  16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.43		0.000	8892	350140	1	
2	1.73		0.000	99946	1139316	1	
3	2.89		0.000	7710	145603	2	
4	3.00		0.000	5008	34222	2	
9	4.17		0.000	3631	30871	2	
10	4.33		0.000	3664	63226	2	
11	4.63		0.000	5149	50021	2	
12	4.79		0.000	5862	41909	2	
13	4.93		0.000	7044	33225	2	
14	5.04		0.000	8013	38580	2	
15	5.15		0.000	9521	178095	2	
16	5.44		0.000	9745	64773	2	
17	5.60		0.000	11079	394306	2	
18	7.17	fluoride	996.232	2712993	42063262	2	0.00
19	7.87	glycolate	1156.356	529287	9544163	2	0.00
20	8.35	acetate	2859.626	1168398	23862149	2	0.00
21	9.04	formate	2630.710	1994591	37969719	2	0.00
22	9.81		0.000	142353	2756509	2	
Totals			7642.925	6732886	118760089		

File: 99022701.D23 Sample: S98T002960 SPK

```

=====
Sample Name: S98T002961      Date: 02/27/1999 21:55:53
Data File  : C:\DX\DATA\99022701.D24
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 24      Detector: CDM-2
Analyst    :                  Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1111    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	2188	33732	1	
3	1.70		0.000	104435	1211700	1	
4	2.95		0.000	10315	232530	2	
5	3.07		0.000	8970	101930	2	
6	3.35		0.000	2845	45018	3	
12	4.17		0.000	2265	19020	2	
13	4.35		0.000	3312	87117	2	
14	4.77		0.000	5639	37507	2	
15	4.92		0.000	27348	1462475	2	
16	7.15	fluoride	316.896	786321	12475871	2	0.00
17	8.29	acetate	1917.621	724340	16006329	2	0.00
18	9.01	formate	1352.524	997777	18895875	2	0.00
19	9.79		0.000	158024	3057014	2	
Totals			3587.041	2833777	53666116		

File: 99022701.D24 Sample: S98T002961

```

=====
Sample Name: S98T002961 DUP                      Date: 02/27/1999 22:16:02
Data File  : C:\DX\DATA\99022701.D25
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 25              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1111    5700 5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	8086	294048	1	
2	1.63		0.000	5481	181608	1	
4	3.07		0.000	18535	513829	2	
5	3.20		0.000	17094	214819	2	
6	3.47		0.000	14251	338800	2	
7	3.87		0.000	8087	66028	2	
8	4.12		0.000	5144	59838	2	
9	4.24		0.000	3603	30685	2	
13	5.03		0.000	4030	28365	2	
14	5.23		0.000	7142	78391	2	
15	5.40		0.000	9204	541411	2	
16	7.17	fluoride	323.661	808748	12767010	2	0.00
17	8.32	acetate	1970.784	749305	16451831	2	0.00
18	9.04	formate	1391.011	1031424	19461552	2	0.00
19	9.81		0.000	163086	3127206	2	
Totals			3685.456	2853220	54155422		

File: 99022701.D25 Sample: S98T002961 DUP

```

=====
Sample Name: S98T002961 SPK                      Date: 02/27/1999 22:36:03
Data File  : C:\DX\DATA\99022701.D26
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 26              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----

```

```

External      1      1111    5700    5Hz    0.00  19.00      16000
-----

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	9214	303626	1	
2	1.73		0.000	104318	1189650	1	
3	2.95		0.000	15825	411817	2	
4	3.16		0.000	13193	197648	2	
5	3.55		0.000	6187	117885	3	
12	4.81		0.000	1492	19079	2	
17	5.53		0.000	3911	19447	2	
18	5.61		0.000	5410	43575	2	
19	5.75		0.000	7055	214788	2	
20	7.17	fluoride	972.067	2673497	40998375	2	0.00
21	7.87	glycolate	1192.278	542006	9853417	2	0.00
22	8.35	acetate	3109.151	1268559	25929702	2	0.00
23	9.04	formate	2795.561	2127368	40474391	2	0.00
24	9.81		0.000	160585	3104906	2	
Totals			8069.057	6938621	122878306		

File: 99022701.D26 Sample: S98T002961 SPK

Sample Name: S98T002962 Date: 02/27/1999 22:57:08
 Data File : C:\DX\DATA\99022701.D27
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 27 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1111	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	1397	17660	1	
2	1.70		0.000	105747	1228274	1	
3	2.97		0.000	21467	476103	2	
4	3.17		0.000	21578	175777	2	
5	3.27		0.000	21338	207128	2	
6	3.64		0.000	22517	349466	2	
7	3.75		0.000	22908	113769	2	
8	3.81		0.000	23045	223069	2	
9	4.15		0.000	23009	492452	2	
10	4.36		0.000	22252	319453	2	
11	4.65		0.000	22082	159774	2	
12	4.77		0.000	22345	197256	2	
13	4.92		0.000	21950	161384	2	
14	5.07		0.000	21695	189324	2	
15	5.21		0.000	21661	340406	2	
16	5.57		0.000	20468	826484	2	
17	7.15	fluoride	233.167	537621	8877976	2	0.00
18	8.32	acetate	2064.406	781664	17235737	2	0.00
19	9.01	formate	1345.630	991880	18794614	2	0.00
20	9.79		0.000	161818	3164486	2	
Totals			3643.203	2888441	53550594		

File: 99022701.D27 Sample: S98T002962

```

=====
Sample Name: S98T002962 DUP                      Date: 02/27/1999 23:18:07
Data File  : C:\DX\DATA\99022701.D28
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 28              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1111    5700  5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
3	1.70		0.000	106010	1217578	1	
4	3.01		0.000	17135	447053	2	
5	3.37		0.000	13745	380442	2	
6	3.64		0.000	10160	89330	2	
7	3.81		0.000	8021	72600	2	
8	4.01		0.000	5817	74736	2	
9	4.21		0.000	3574	32050	2	
10	4.36		0.000	2859	32689	2	
13	4.81		0.000	2927	30150	2	
14	5.12		0.000	7152	78895	2	
15	5.23		0.000	8697	70831	2	
16	5.40		0.000	10355	82996	2	
17	5.51		0.000	11801	513177	2	
18	7.15	fluoride	236.989	553631	9041969	2	0.00
19	8.32	acetate	2124.065	808626	17734850	2	0.00
20	9.01	formate	1387.320	1031378	19407276	2	0.00
21	9.79		0.000	167221	3258550	2	
Totals			3748.373	2769108	52565172		

File: 99022701.D28 Sample: S98T002962 DUP

```

=====
Sample Name: S98T002962 SPK                      Date: 02/27/1999 23:37:57
Data File  : C:\DX\DATA\99022701.D29
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 29              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          1111    5700  5Hz   0.00 19.00    16000
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	1.70		0.000	108160	1267502	1	
3	2.95		0.000	22234	407299	2	
4	3.12		0.000	22495	396928	2	
5	3.37		0.000	21943	282581	2	
6	3.56		0.000	22589	616448	2	
7	4.17		0.000	21564	237838	2	
8	4.28		0.000	21521	275484	2	
9	4.51		0.000	21520	241104	2	
10	4.71		0.000	21070	235765	2	
11	5.01		0.000	20778	555979	2	
12	5.36		0.000	20199	145484	2	
13	5.55		0.000	20865	811308	2	
14	7.15	fluoride	863.251	2338988	36214758	2	0.00
15	7.84	glycolate	1217.759	555428	10073068	2	0.00
16	8.32	acetate	3282.983	1331255	27366811	2	0.00
17	9.01	formate	2839.242	2145595	41139839	2	0.00
18	9.79		0.000	168889	3338510	2	
Totals			8203.236	6885092	123606706		

File: 99022701.D29 Sample: S98T002962 SPK

LABCORE Completed Worklist Report for Worklist# 28406

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 rerun @ic4g-01

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	2.96e+01	0.296	ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	7.90e-02	0.079	ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	5.85e+01	95.902 %	Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.08e+02	93.103 %	Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.18e+02	99.160 %	Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.08e2	9.91e+01	91.759 %	Recovery
3 CCV	0	@IC4G-QC F*4	QC	6.29e1	5.68e+01	90.302 %	Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.16e+02	96.667 %	Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.21e+02	100.833 %	Recovery
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.03e2	9.39e+01	91.165 %	Recovery
4 BLNK-PREP	0	@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 ACETATE2	SOLID	1	3.20e-02	0.032	ug/g
4 BLNK-PREP	0	@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	@IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002848	0 W @IC4G-01 F*4-01	SOLID	N/A	8.628e+02	127.400	ug/g
5 SAMPLE	S98T002848	0 W @IC4G-01 ACETATE2	SOLID	N/A	6.609e+02	82.220	ug/g
5 SAMPLE	S98T002848	0 W @IC4G-01 GLYCOLT1	SOLID	N/A	2.302e+02	230.200	ug/g
6 DUP	S98T002848	0 W @IC4G-01 F*4-01	SOLID	8.63e+02	1.61e+03	60.412	RPD
6 DUP	S98T002848	0 W @IC4G-01 ACETATE2	SOLID	6.61e+02	6.77e+02	2.392	RPD
6 DUP	S98T002848	0 W @IC4G-01 GLYCOLT1	SOLID	<2.30e2			RPD
7 SPK	S98T002848	0 W @IC4G-01 F*4-01	SOLID	6.10e1	6.24e+01	102.295 %	Recovery
7 SPK	S98T002848	0 W @IC4G-01 ACETATE2	SOLID	1.16e2	1.06e+02	91.379 %	Recovery
7 SPK	S98T002848	0 W @IC4G-01 GLYCOLT1	SOLID	1.08e2	1.02e+02	94.444 %	Recovery

Final page for worklist# 28406

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 28406

Analyst: ADP Instrument: IC _____Book# 59N21A
58N21AMethod: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 rerun @ic4g-01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002848	0 W	@IC4G-01	SOLID	98000437	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, GLYCOLT1						
6 DUP	S98T002848	0 W	@IC4G-01	SOLID		
7 SPK	S98T002848	0 W	@IC4G-01	SOLID		

Final page for worklist # 28406

Signature _____ Date _____

28406 Feb. cs v

Signature _____ Date _____

Data Entry Comments:

Uploaded & Val. data 3/6/99 JM Fyke

Sample Name: 59N21-LCS Date: 02/18/1999 12:47:46
 Data File : E:\DATA\99021801.d30
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 30 Detector:CDM-2
 Analyst : JMF for ADP Column: AG14A-SC,AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject
 External 1 101 4557 5Hz 0.00 15.19 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	2.23		0.000	4926	91607	2	
3	2.68		0.000	3067	38800	2	
4	2.93		0.000	4417	38284	2	
5	3.04		0.000	6153	156164	3	
7	3.89		0.000	1658	17594	2	
9	4.27		0.000	2643	34647	2	
10	4.52		0.000	2852	29513	2	
11	4.88		0.000	1333	20190	2	
12	5.24		0.000	2947	39283	2	
13	5.56		0.000	2288	37744	2	
17	7.33	fluoride	58.473	1788012	26598552	2	0.00
18	8.05	glycolate	99.138	482593	8978537	2	0.00
19	8.56	acetate	107.784	503064	9845652	2	0.00
20	9.28	formate	117.560	910545	18024310	2	0.00
Totals			382.954	3716498	63950878		

File: 99021801.d30 Sample: 59N21-LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 1762 TO 1787.

```

=====
Sample Name: CCV                               Date: 02/18/1999 14:00:10
Data File  : E:\DATA\99021801.d31
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 31           Detector: CDM-2
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700   5Hz   0.00  19.00      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	14943	371777	2	
2	1.57		0.000	10834	247804	2	
4	2.84		0.000	17739	223677	2	
5	2.91		0.000	20041	132070	2	
6	3.09		0.000	25852	205900	2	
13	5.69		0.000	2099	20231	2	
16	6.09		0.000	1671	20183	2	
20	7.44	fluoride	56.837	1739739	25815188	2	0.00
21	8.13	glycolate	93.889	459489	8483819	2	0.00
22	8.64	acetate	116.028	537386	10611581	2	0.00
23	9.36	formate	121.215	977515	18614437	2	0.00
Totals			387.968	3807309	64746667		

File: 99021801.d31 Sample: CCV *58N21A*

Data Reprocessed On 02/26/1999 15:32:23

Sample Name: PREP BLK	Date: 02/18/1999 14:26:15
Data File : E:\DATA\99021801.d32	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 32	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	3722	60128	1	
3	2.27		0.000	2249	23423	2	
7	3.15		0.000	2961	22015	2	
9	3.45		0.000	2368	35086	2	
12	4.31		0.000	3234	69846	2	
13	4.68		0.000	3177	23957	2	
17	5.36		0.000	3229	31719	2	
18	5.57		0.000	2480	32834	2	
20	6.35		0.000	3971	42659	2	
25	7.41	fluoride	0.024	3292	65783	2	0.00
28	8.45	acetate	0.032	2894	63767	2	0.00
29	8.85		0.000	3027	34797	2	
30	9.39	formate	0.047	3427	78603	2	0.00
31	9.84		0.000	2703	29262	2	
Totals			0.103	42734	613880		

File: 99021801.d32 Sample: PREP BLK

1784

Data Reprocessed On 02/26/1999 15:32:24

Sample Name: S98T002848	Date: 02/18/1999 14:47:05
Data File : E:\DATA\99021801.d33	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 33	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	11252	380964	1	
2	1.73		0.000	10045	82928	1	
3	2.10		0.000	1208	16924	1	
4	2.93		0.000	7406	159957	2	
5	3.11		0.000	7500	104696	2	
6	3.33		0.000	6204	35525	2	
7	3.48		0.000	5544	89027	2	
8	3.77		0.000	4863	78427	2	
9	4.05		0.000	4456	36363	2	
10	4.32		0.000	4121	50190	2	
11	4.48		0.000	4248	23640	2	
12	4.55		0.000	4304	39875	2	
13	4.73		0.000	2966	17588	2	
14	4.88		0.000	1986	17824	2	
21	6.03		0.000	3311	50022	2	
22	7.36	fluoride	4.407	556396	8877508	2	0.00
23	8.59	acetate	3.376	56568	1287426	2	0.00
24	9.31	formate	2.066	45368	895755	2	0.00
Totals			9.849	737747	12244640		

File: 99021801.d33 Sample: S98T002848

Data Reprocessed On 02/26/1999 15:32:25

```

=====
Sample Name: S98T002848 DUP                      Date: 02/18/1999 15:07:57
Data File  : E:\DATA\99021801.d34
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 34              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21  5700  5Hz  0.00 19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	10133	345504	1	
3	1.73		0.000	12203	120652	1	
4	2.21		0.000	2757	32662	1	
5	3.00		0.000	7462	202310	2	
6	3.21		0.000	6010	42040	2	
7	3.40		0.000	5372	83142	2	
8	3.67		0.000	3371	23816	2	
9	3.85		0.000	3373	27899	2	
10	4.03		0.000	2612	38367	2	
12	4.39		0.000	2812	18530	2	
17	5.03		0.000	2144	23591	2	
18	5.25		0.000	3170	45990	2	
19	5.56		0.000	4601	33878	2	
20	5.64		0.000	5166	124972	2	
21	6.11		0.000	4287	55269	2	
22	7.36	fluoride	8.213	1129628	17547253	2	0.00
23	8.56	acetate	3.458	57627	1324732	2	0.00
24	9.31	formate	2.215	51508	1008190	2	0.00

```

-----
Totals                      13.886      1314235      21098798
-----

```

File: 99021801.d34 Sample: S98T002848 DUP

Data Reprocessed On 02/26/1999 15:32:25

```

=====
Sample Name: S98T002848 SPK                      Date: 02/18/1999 15:30:56
Data File  : E:\DATA\99021801.d35
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 35              Detector: CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700    5Hz    0.00  19.00      16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	10152	348356	2	
2	1.73		0.000	13056	118364	2	
4	2.77		0.000	5150	63463	2	
5	2.91		0.000	4751	35361	2	
6	3.12		0.000	4310	49247	2	
7	3.27		0.000	4058	52142	2	
8	3.55		0.000	2507	34561	2	
15	4.87		0.000	2588	33566	2	
16	5.12		0.000	2622	19462	2	
18	5.39		0.000	3226	29691	2	
19	5.60		0.000	4109	46921	2	
20	5.77		0.000	5302	159490	2	
22	7.39	fluoride	16.887	2498983	37538382	2	0.00
23	8.08	glycolate	20.305	472833	8838730	2	0.00
24	8.59	acetate	24.632	552712	10838157	2	0.00
25	9.31	formate	25.085	962929	18522583	2	0.00
Totals			86.908	4549287	76728477		

File: 99021801.d35 Sample: S98T002848 SPK

1787

LABCORE Completed Worklist Report for Worklist# 28476

Analyst: ehc

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY105 for @IC4G-01

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0	IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0	IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0	IC4G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0	IC4G-QC F*4	QC	6.10e1	5.93e+01	97.213 % Recovery	
2 LCS-INST	0	IC4G-QC ACETATE2	QC	1.16e2	1.09e+02	93.966 % Recovery	
2 LCS-INST	0	IC4G-QC FORMATE2	QC	1.19e2	1.21e+02	101.681 % Recovery	
2 LCS-INST	0	IC4G-QC GLYCOLT1	QC	1.08e2	1.01e+02	93.519 % Recovery	
3 CCV	0	IC4G-QC F*4	QC	6.29e1	5.48e+01	87.122 % Recovery	
3 CCV	0	IC4G-QC ACETATE2	QC	1.20e2	1.13e+02	94.167 % Recovery	
3 CCV	0	IC4G-QC FORMATE2	QC	1.20e2	1.20e+02	100.000 % Recovery	
3 CCV	0	IC4G-QC GLYCOLT1	QC	1.03e2	9.33e+01	90.583 % Recovery	
4 BLNK-PREP	0	IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0	IC4G-01 ACETATE2	SOLID	1	2.90e-02	0.029	ug/g
4 BLNK-PREP	0	IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0	IC4G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002614	0 W IC4G-01 ACETATE2	SOLID	N/A	1.408e+03	23.420	ug/g
5 SAMPLE	S98T002614	0 W IC4G-01 FORMATE2	SOLID	N/A	2.843e+02	72.600	ug/g
5 SAMPLE	S98T002614	0 W IC4G-01 GLYCOLT1	SOLID	N/A	6.557e+01	65.580	ug/g
6 DUP	S98T002614	0 W IC4G-01 F*4-01	SOLID	1.40e+02	2.26e+02	46.995	RPD
6 DUP	S98T002614	0 W IC4G-01 ACETATE2	SOLID	1.41e+03	1.49e+03	5.517	RPD
6 DUP	S98T002614	0 W IC4G-01 FORMATE2	SOLID	2.84e+02	3.30e+02	14.984	RPD
6 DUP	S98T002614	0 W IC4G-01 GLYCOLT1	SOLID	<6.56e1	<6.76e1		RPD
7 SPK	S98T002614	0 W IC4G-01 F*4-01	SOLID	6.10e1	6.11e+01	100.164 % Recovery	
7 SPK	S98T002614	0 W IC4G-01 ACETATE2	SOLID	1.16e2	1.38e+02	118.966 % Recovery	
7 SPK	S98T002614	0 W IC4G-01 FORMATE2	SOLID	1.19e2	1.33e+02	111.765 % Recovery	
7 SPK	S98T002614	0 W IC4G-01 GLYCOLT1	SOLID	1.08e2	1.08e+02	100.000 % Recovery	
8 SAMPLE	S98T002615	0 W IC4G-01 ACETATE2	SOLID	N/A	1.892e+03	24.020	ug/g
8 SAMPLE	S98T002615	0 W IC4G-01 FORMATE2	SOLID	N/A	3.602e+02	74.460	ug/g
8 SAMPLE	S98T002615	0 W IC4G-01 GLYCOLT1	SOLID	N/A	6.727e+01	67.260	ug/g
9 DUP	S98T002615	0 W IC4G-01 F*4-01	SOLID	7.21e+01	6.20e+01	15.063	RPD
9 DUP	S98T002615	0 W IC4G-01 ACETATE2	SOLID	1.89e+03	1.90e+03	0.528	RPD
9 DUP	S98T002615	0 W IC4G-01 FORMATE2	SOLID	3.60e+02	3.53e+02	1.954	RPD
9 DUP	S98T002615	0 W IC4G-01 GLYCOLT1	SOLID	<6.73e1	<6.53e1		RPD

Final page for worklist# 28476

Analyst Signature Date

Analyst Signature Date

Jane Lye 3/12/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 28476

Analyst: EMC Instrument: IC 4552 Book# 59N21-A
58N21-A
 Method: LA-533-105 Rev/Mod F-0
 Worklist Comment: BY105 for @IC4G-01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002614	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
6 DUP	S98T002614	0 W	@IC4G-01	SOLID		
7 SPK	S98T002614	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T002615	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
9 DUP	S98T002615	0 W	@IC4G-01	SOLID		

Final page for worklist # 28476

EMC 3-9-99
 Signature Date

_____
 Signature Date

Data Entry Comments:

uploaded 3-9-99
John Howell
28476MAR.CSV
Rejected F due to std failure
Rest validated 3/12/99 JM Long
Reanalyzed on W/L# 28541.
3/24/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

Sample Name: INSTR BLANK Date: 03/09/1999 08:40:51
 Data File : C:\DX\DATA\99030911.D08
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 8 Detector: CDM-2
 Analyst : *[Signature]* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	4848	5Hz	0.00	16.16	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.47		0.000	148442	3084057	2	
2	1.17		0.000	683521	14852166	2	
3	2.27		0.000	107998	687180	2	
4	2.51		0.000	655667	53574213	2	
5	3.23		0.000	1915474	83755336	2	
6	5.79		0.000	14476178	1341278258	2	

Totals 0.000 17987279 1497231210

File: 99030911.D08 Sample: INSTR BLANK

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1790 TO 1799.

Sample Name: 59N21-A LCS Date: 03/09/1999 08:58:59
 Data File : C:\DX\DATA\99030911.D09
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 9 Detector: CDM-2
 Analyst : *CAE* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	4994	5Hz	0.00	16.64	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	14163	181764	1	
4	2.51		0.000	2487	105121	2	
5	3.11		0.000	5360	35315	2	
6	3.25		0.000	4361	21197	2	
7	3.44		0.000	4263	71859	2	
8	3.97		0.000	2906	40061	2	
9	4.32		0.000	3408	46552	2	
10	4.68		0.000	3580	18664	1	
11	4.91		0.000	3727	33884	2	
13	5.37		0.000	2219	21539	2	
14	5.59		0.000	2734	18137	2	
15	5.75		0.000	3890	26591	2	
16	5.93		0.000	4925	55317	2	
18	6.35		0.000	2568	24492	2	
19	7.12	fluoride	59.332	1785016	27009642	2	0.00
20	7.81	glycolate	101.213	510375	9174494	2	0.00
21	8.29	acetate	109.433	504970	9998948	2	0.00
22	8.99	formate	120.818	980229	18550279	2	0.00
Totals			390.796	3841181	65433856		

File: 99030911.D09 Sample: 59N21-A LCS

```

=====
Sample Name: 58N21-A CCV                      Date: 03/09/1999 09:49:45
Data File  : C:\DX\DATA\99030911.D12
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 12          Detector: CDM-2
Analyst    : PN Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101  4413  5Hz   0.00 14.71 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
2	2.13		0.000	10171	227748	2	
3	2.25		0.000	2773	17262	2	
4	2.57		0.000	10310	238134	2	
5	2.97		0.000	8915	87549	2	
6	3.21		0.000	6478	136965	2	
7	3.83		0.000	3099	40148	2	
11	4.65		0.000	3329	32152	2	
17	5.76		0.000	7115	79146	2	
18	5.92		0.000	4475	38938	2	
19	6.16		0.000	2485	16054	2	
21	6.39		0.000	3276	29315	2	
22	7.12	fluoride	54.829	1665196	24854621	2	0.00
23	7.81	glycolate	93.296	467609	8428023	2	0.00
24	8.29	acetate	113.338	525856	10361775	2	0.00
25	8.99	formate	120.122	950968	18438012	2	0.00
Totals			381.585	3672054	63025840		

File: 99030911.D12 Sample: 58N21-A CCV

```

=====
Sample Name: PREP BLANK                      Date: 03/09/1999 09:33:55
Data File  : C:\DX\DATA\99030911.D11
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 11         Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====

```

```

External      1      1      4550 5Hz      0.00 15.16      16000
=====

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	28175	401622	1	
2	1.00		0.000	5516	104596	1	
3	2.19		0.000	2111	19231	2	
5	2.59		0.000	3341	39328	2	
6	2.95		0.000	3370	26926	2	
7	3.17		0.000	2636	22128	2	
9	3.76		0.000	8828	112541	2	
10	4.01		0.000	7267	53945	2	
11	4.17		0.000	5566	125162	2	
12	4.52		0.000	6224	33648	2	
13	4.61		0.000	8523	131285	2	
15	5.11		0.000	2321	19151	2	
16	5.39		0.000	4091	26859	2	
17	5.49		0.000	5057	25417	2	
18	5.64		0.000	5831	47266	2	
19	5.83		0.000	4415	31293	2	
20	5.99		0.000	3153	25341	2	
21	6.27		0.000	3599	30676	2	
22	6.53		0.000	2962	30827	2	
23	6.80		0.000	1969	28704	2	
24	7.12	fluoride	0.030	12027	354748	2	0.00
26	8.21		0.000	3498	31311	2	
27	8.43	acetate	0.029	3698	37920	2	0.00
28	8.69		0.000	2593	30399	2	
29	9.09	formate	0.043	2191	23344	1	0.00
32	9.79		0.000	2076	30276	1	
Totals			0.103	141038	1843945		

File: 99030911.D11 Sample: PREP BLANK

Sample Name: S98T002614 Date: 03/09/1999 12:06:37
 Data File : C:\DX\DATA\99030911.D14
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 14 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	4546	5Hz	0.00	15.15	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.47		0.000	23791	606421	2	
2	1.73		0.000	46670	349462	2	
4	2.76		0.000	4165	33124	2	
5	2.84		0.000	4149	24381	2	
6	2.96		0.000	3925	44023	2	
7	3.23		0.000	3072	22508	2	
9	3.48		0.000	3621	27594	2	
10	3.72		0.000	2573	22802	2	
11	3.97		0.000	2123	35260	2	
12	4.36		0.000	2645	16464	2	
21	5.80		0.000	2133	52060	1	
22	7.15	fluoride	0.716	274438	4565890	1	0.00
23	8.29	acetate	7.215	572849	11114636	2	0.00
24	8.96	formate	1.457	184128	3190125	2	0.00
Totals			9.387	1130284	20104750		

File: 99030911.D14 Sample: S98T002614

```

=====
Sample Name: S98T002614 DUP                      Date: 03/09/1999 12:22:35
Data File  : C:\DX\DATA\99030911.D15
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 15              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1           6  4548  5Hz  0.00 15.16   16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.70		0.000	23610	231300	1	
2	2.76		0.000	15722	282059	2	
3	2.93		0.000	15962	102691	2	
4	3.07		0.000	15195	136466	2	
5	3.23		0.000	14646	95091	2	
6	3.33		0.000	16260	793728	2	
7	4.63		0.000	10568	79515	2	
8	4.75		0.000	11591	264836	2	
9	5.19		0.000	10677	162638	2	
10	5.48		0.000	11400	399421	2	
11	7.09	fluoride	1.122	486673	7787155	2	0.00
12	8.24	acetate	7.385	537573	11380480	2	0.00
13	8.96	formate	1.640	188486	3676434	2	0.00
14	9.73		0.000	21974	413480	2	
Totals			10.147	1380339	25805295		

File: 99030911.D15 Sample: S98T002614 DUP

```

=====
Sample Name: S98T002614 SPK                      Date: 03/09/1999 12:38:08
Data File  : C:\DX\DATA\99030911.D16
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 16              Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 4427 5Hz 0.00 14.75 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.70		0.000	28508	281356	1	
2	2.75		0.000	10336	110188	2	
3	2.81		0.000	11513	312355	2	
4	3.31		0.000	10498	57380	2	
5	3.44		0.000	11771	370703	2	
6	4.01		0.000	11668	372649	2	
7	4.56		0.000	11394	476928	2	
8	5.31		0.000	15250	724866	2	
9	7.09	fluoride	3.772	1877913	29003521	2	0.00
10	7.79	glycolate	5.419	452783	8232031	2	0.00
11	8.27	acetate	14.100	1070009	21794925	2	0.00
12	8.96	formate	8.091	1082179	21038208	3	0.00
13	9.76		0.000	22298	393481	4	
Totals			31.381	4616120	83168593		

File: 99030911.D16 Sample: S98T002614 SPK

=====

Sample Name: S98T002615	Date: 03/09/1999 12:53:28
Data File : C:\DX\DATA\99030911.D17	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 17	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	6	4372	5Hz	0.00	14.57	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	1.70		0.000	29135	282600	1	
2	2.75		0.000	957	24480	1	
7	4.09		0.000	525	16580	1	
9	4.81		0.000	2924	52027	2	
11	7.09	fluoride	0.360	93434	1752798	2	0.00
12	8.24	acetate	9.450	678318	14597980	2	0.00
13	8.96	formate	1.799	207589	4097758	2	0.00
14	9.71		0.000	11253	222439	2	
Totals			11.608	1024135	21046662		

File: 99030911.D17 Sample: S98T002615

Sample Name: S98T002615 DUP Date: 03/09/1999 13:05:49
 Data File : C:\DX\DATA\99030911.D18
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 18 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 3469 5Hz 0.00 11.56 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.50		0.000	6388	45972	1	
3	1.73		0.000	26605	236590	1	
4	2.79		0.000	9260	112150	2	
5	2.92		0.000	11019	524446	3	
13	5.55		0.000	2490	25740	2	
14	7.12	fluoride	0.319	79963	1427109	2	0.00
15	8.27	acetate	9.794	710689	15132528	2	0.00
16	8.99	formate	1.820	215259	4154834	2	0.00
17	9.76		0.000	13133	262484	2	
Totals			11.933	1074806	21921854		

File: 99030911.D18 Sample: S98T002615 DUP

LABCORE Completed Worklist Report for Worklist# 28541

Analyst: rdm

Instrument: IC45S2

Book#: 58N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 RERUN DF=21 JHW

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	<2.10e-2		ug/mL
1 CCB	0		@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0		@IC4G-QC F*4	QC	5.10e1	6.22e+01	101.967 %	Recovery
2 LCS-INST	0		@IC4G-QC ACETATE2	QC	1.16e2	1.13e+02	97.414 %	Recovery
2 LCS-INST	0		@IC4G-QC FORMATE2	QC	1.13e2	1.23e+02	103.361 %	Recovery
2 LCS-INST	0		@IC4G-QC GLYCOLTI	QC	1.08e2	1.04e+02	96.296 %	Recovery
3 CCV	0		@IC4G-QC F*4	QC	6.29e1	5.95e+01	94.595 %	Recovery
3 CCV	0		@IC4G-QC ACETATE2	QC	1.20e2	1.21e+02	100.833 %	Recovery
3 CCV	0		@IC4G-QC FORMATE2	QC	1.20e2	1.26e+02	105.000 %	Recovery
3 CCV	0		@IC4G-QC GLYCOLTI	QC	1.03e2	9.92e+01	96.311 %	Recovery
4 BLNK-PREP	0		@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002932	0 W	@IC4G-01 F*4-01	SOLID	N/A	2.298e+03	625.700	ug/g
6 DUP	S98T002932	0 W	@IC4G-01 F*4-01	SOLID	2.30e+03	2.26e+03	1.754	RPD
7 SPK	S98T002932	0 W	@IC4G-01 F*4-01	SOLID	6.10e1	5.80e+01	95.082 %	Recovery

Final page for worklist# 28541

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 3/12/99

LABCORE Data Entry Template for Worklist# 28541

Analyst: RDM Instrument: IC 4552 Book# 59U21A
58N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 RERUN DF=21 JHW

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4-QC	QC		
2	ICV				@IC4-QC	QC		
3	CCV				@IC4-QC	QC		
4	BLNK-PREP				@IC4G-01	SOLID		
5	SAMPLE	S98T002932	0	W	@IC4G-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: F*4-01						
6	DUP	S98T002932	0	W	@IC4G-01	SOLID		
7	SPK	S98T002932	0	W	@IC4G-01	SOLID		

WRONG TRAY fixed 3/12/99 MTJ

Final page for worklist # 28541

See Original
Signature _____ Date _____

Signature _____ Date _____

Data Entry Comments:

uploaded 3-9-99 re uploaded to fix tray 3/12/99 MTJ
JR Howell Validated 3/12/99 MTJ
28541/MAR.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 28447

Analyst: Rom Instrument: IC 4552 Book# 54N21A
54N21A

Method: LA-533-105 Rev/Mod P-0

Worklist Comment: BY-105 RERUN DF=21 JHW

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC-QC	QC		
2 LCS-INST			@IC-QC	QC		
3 CCV			@IC-QC	QC		
4 BLNK-PREP			@IC-01	SOLID		
5 SAMPLE	S98T002932 0 W		@IC4G-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: F*4-01					
6 DUP	S98T002932 0 W		@IC4G-01	SOLID		
7 SPK	S98T002932 0 W		@IC4G-01	SOLID		

-Wrong Tray

Final page for worklist # 28447

Rom
Signature

3/4/99
Date

_____
Signature

_____
Date

Data Entry Comments: 54N21A used as spike

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 03/04/1999 19:15:07
 Data File : C:\DX\DATA\99030311.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst : *JMF for RDM* Column: AG14A-SC, AS14A-SC, SRS
3/5/99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	1	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	29268	361192	1	
2	2.03		0.000	28511	389560	1	
3	2.47		0.000	8290	66963	2	
8	3.07		0.000	15817	54001	2	
9	3.15		0.000	43052	305343	2	
10	3.31		0.000	96922	9563668	2	
11	4.11		0.000	321871	1801987	2	
12	4.21		0.000	352784	19295538	2	
13	6.88		0.000	2336412	222999085	2	
Totals			0.000	3232928	254837338		

File: 99030311.D01 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1803 TO 1809.

Sample Name: LCS 59N21-A Date: 03/04/1999 19:39:14
Data File : C:\DX\DATA\99030311.D02
Method : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2
Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	20789	837568	1	
9	3.33		0.000	1692	19503	2	
12	4.04		0.000	2098	16273	2	
28	7.23	fluoride	62.166	1936791	28368028	2	0.00
29	7.89	glycolate	103.854	518101	9424069	2	0.00
30	8.37	acetate	112.804	528412	10312153	2	0.00
31	9.07	formate	123.095	999851	18918223	2	0.00
Totals			401.919	4007734	67895816		

File: 99030311.D02 Sample: LCS 59N21-A

```

=====
Sample Name: CCV 58N21-A                      Date: 03/04/1999 20:01:33
Data File  : C:\DX\DATA\99030311.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3
Analyst    :                               Detector: CDM-2
                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700  5Hz   0.00  19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	1956	56324	1	
8	2.89		0.000	2088	20321	2	
9	3.05		0.000	2218	24381	2	
15	4.20		0.000	1301	18206	1	
18	4.99		0.000	1753	26800	2	
20	5.44		0.000	1833	23668	2	
25	7.20	fluoride	59.532	1852898	27105584	2	0.00
26	7.89	glycolate	99.187	502947	8983147	2	0.00
27	8.35	acetate	120.937	560459	11067269	2	0.00
28	9.07	formate	126.459	1048067	19462088	2	0.00
Totals			406.114	3975521	66787790		

File: 99030311.D03 Sample: CCV 58N21-A

1805

```
=====
Sample Name: PREP BLK                      Date: 03/04/1999 20:27:47
Data File  : C:\DX\DATA\99030311.D04
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 4
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====
```

```
-----
Calibration Volume  Dilution Points Rate  Start  Stop Area Reject
-----
External           1           1    5700  5Hz   0.00  19.00    16000
-----
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	9977	584814	1	
20	5.60		0.000	1492	16170	2	
28	7.23	fluoride	0.024	924	48849	2	0.00
Totals			0.024	12393	649833		

```

=====
Sample Name: S98T002932      Date: 03/04/1999 20:55:29
Data File  : C:\DX\DATA\99030311.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5      Detector: CDM-2
Analyst    :                  Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700    5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	13128	703944	1	
2	2.25		0.000	3559	47360	1	
4	2.64		0.000	2291	24183	2	
5	2.97		0.000	2142	41941	3	
17	5.17		0.000	1123	17904	2	
21	6.09		0.000	2137	25161	2	
23	7.20	fluoride	11.497	285806	4308063	2	0.00
24	8.48	acetate	4.953	8896	225974	2	0.00
25	9.04	formate	5.539	5556	205364	2	0.00
Totals			21.989	324638	5599895		

File: 99030311.D05 Sample: S98T002932

```

=====
Sample Name: S98T002932DUP      Date: 03/04/1999 21:16:59
Data File  : C:\DX\DATA\99030311.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6      Detector: CDM-2
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101  5700  5Hz  0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	9505	331396	1	
3	2.23		0.000	3471	49534	1	
4	2.73		0.000	3320	45488	2	
5	3.08		0.000	4101	124619	2	
6	3.61		0.000	1978	35633	2	
15	5.75		0.000	1569	17749	2	
18	7.20	fluoride	11.149	257507	4144222	3	0.00
20	8.40	acetate	3.112	2800	52686	4	0.00
22	9.09	formate	4.588	2857	56264	2	0.00
23	9.84		0.000	1498	17513	2	
Totals			18.849	288607	4875105		

File: 99030311.D06 Sample: S98T002932DUP

```

=====
Sample Name: S98T002932SPK      Date: 03/04/1999 21:37:47
Data File : C:\DX\DATA\99030321.D07
Method    : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 7      Detector: CDM-2
Analyst   :      Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----

```

```

External      1      101      5700      5Hz      0.00      19.00      16000
-----

```

```

***** Peak Report: All Peaks *****

```

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.47		0.000	1291	24204	1	
3	2.27		0.000	2537	28064	1	
8	3.31		0.000	1452	19912	1	
11	4.39		0.000	1526	18650	2	
22	7.17	fluoride	69.468	2135899	31874434	2	0.00
23	7.87	glycolate	98.868	490475	8953067	2	0.00
24	8.35	acetate	106.239	506359	9702080	2	0.00
25	9.04	formate	115.475	936635	17688077	2	0.00
Totals			390.051	4076175	68308488		

File: 99030321.D07 Sample: S98T002932SPK

LABCORE Completed Worklist Report for Worklist# 28542

Analyst: krm

Instrument: IC45S2

Book#: 58N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 rerun for F. Use 21 df. jmf

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0		@IC4G-QC F*4	QC	1	<1.10e-2		ug/mL
1 CCB	0		@IC4G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1 CCB	0		@IC4G-QC FORMATE2	QC	1	4.30e-2	0.043	ug/mL
1 CCB	0		@IC4G-QC GLYCOLTI	QC	1	<2.00e-2		ug/mL
2 LCS-INST	0		@IC4G-QC F*4	QC	6.10e1	6.38e+01	104.590 %	Recovery
2 LCS-INST	0		@IC4G-QC ACETATE2	QC	1.16e2	1.17e+02	100.862 %	Recovery
2 LCS-INST	0		@IC4G-QC FORMATE2	QC	1.19e2	1.27e+02	106.723 %	Recovery
2 LCS-INST	0		@IC4G-QC GLYCOLTI	QC	1.08e2	1.06e+02	98.148 %	Recovery
3 CCV	0		@IC4G-QC F*4	QC	6.29e1	5.88e+01	93.482 %	Recovery
3 CCV	0		@IC4G-QC ACETATE2	QC	1.20e2	1.21e+02	100.833 %	Recovery
3 CCV	0		@IC4G-QC FORMATE2	QC	1.20e2	1.28e+02	106.667 %	Recovery
3 CCV	0		@IC4G-QC GLYCOLTI	QC	1.03e2	1.00e+02	97.087 %	Recovery
4 BLNK-PREP	0		@IC4G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4 BLNK-PREP	0		@IC4G-01 GLYCOLTI	SOLID	1	<5.60e-2		ug/g
5 SAMPLE	S98T002613	0 W	@IC4G-01 F*4-01	SOLID	N/A	5.927e+02	131.600	ug/g
6 DUP	S98T002613	0 W	@IC4G-01 F*4-01	SOLID	5.93e+02	5.77e+02	2.735	RPD
6 DUP	S98T002613	0 W	@IC4G-01 ACETATE2	SOLID	?	2.24e+02		RPD
6 DUP	S98T002613	0 W	@IC4G-01 FORMATE2	SOLID	?	<2.58e2		RPD
6 DUP	S98T002613	0 W	@IC4G-01 GLYCOLTI	SOLID	?	<2.33e2		RPD

Final page for worklist# 28542

Analyst Signature

Date

Analyst Signature

Date

James M. Lyle
Reviewer Signature Date 3/22/99

LABCORE Data Entry Template for Worklist# 28542

Analyst: KRM Instrument: IC 4552 Book# 59N21A
58N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: BY-105 rerun for F. Use 21 df. jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002613	0 W	@IC4G-01	SOLED	98000436	BY-105 (2)
	Analytes Requested: F*4-01					
6 DUP	S98T002613	0 W	@IC4G-01	SOLID		

Final page for worklist # 28542

[Signature] 3-9-99
Signature Date

_____
Signature Date

Data Entry Comments:

updated 3-18-99

Validated 3/22/99 JmFye

J. L. Howell

28542MAR.CSV

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Sample Name: CCB Date: 03/09/1999 20:25:33
 Data File : C:\DX\DATA\99030911.D01
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 1 Detector: CDM-2
 Analyst *W. J. Miller* Column: AG14A-SC, AS14A-SC, SRS

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 1 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	34882	1146536	1	
3	2.32		0.000	1903	16248	2	
5	2.65		0.000	1557	35202	2	
6	3.11		0.000	1584	40859	2	
7	3.49		0.000	1873	16986	2	
10	4.31		0.000	2154	32874	2	
14	5.33		0.000	3366	28778	2	
15	5.55		0.000	6203	59892	2	
16	5.67		0.000	7668	51301	2	
17	5.81		0.000	9229	61661	2	
18	5.92		0.000	9507	257488	2	
19	6.40		0.000	6370	119313	3	
25	8.56		0.000	2320	27843	2	
26	9.12	formate	0.043	991	17623	2	0.00
Totals			0.043	89607	1912603		

File: 99030911.D01 Sample: CCB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1812 TO 1817.


```

=====
Sample Name: LCS 59N21A                      Date: 03/09/1999 23:33:18
Data File  : C:\DX\DATA\99030911.D04
Method      : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 4          Detector: CDM-2
Analyst     :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700 5Hz   0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	12689	648136	1	
4	2.84		0.000	2537	16964	2	
5	2.92		0.000	2342	20183	2	
6	3.11		0.000	2427	33457	2	
7	3.45		0.000	2102	18090	2	
8	3.60		0.000	1864	19035	2	
16	5.31		0.000	1112	22159	2	
18	5.83		0.000	2087	19763	2	
19	6.00		0.000	2211	28321	2	
21	7.25	fluoride	104.5 63.767	1994691	29135847	2	0.00
22	7.95	glycolate	97.9 105.774	539232	9605785	2	0.00
23	8.43	acetate	101.3 117.483	551353	10746696	2	0.00
24	9.12	formate	107.0 127.318	1044494	19601022	2	0.00
Totals			414.342	4159142	69915457		

File: 99030911.D04 Sample: LCS 59N21A

```

=====
Sample Name: CCV 58N21A                      Date: 03/09/1999 23:11:39
Data File  : C:\DX\DATA\99030911.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====
  
```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1          101    5700  5Hz   0.00 19.00    16000
=====
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	5806	125740	1	
23	5.93		0.000	2358	34414	2	
27	7.23	fluoride	93.5 58.838	1837897	26773232	2	0.00
28	7.89	glycolate	97.2 100.167	489096	9075733	2	0.00
29	8.40	acetate	101.2 121.451	571564	11114973	2	0.00
30	9.09	formate	106.8 128.102	1045935	19727985	2	0.00
Totals			408.558	3952656	66852078		

File: 99030911.D03 Sample: CCV 58N21A

```

=====
Sample Name: PREP BLK                      Date: 03/09/1999 23:56:16
Data File  : C:\DX\DATA\99030911.D05
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 5        Detector: CDM-2
Analyst    :                             Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1           1 5700 5Hz 0.00 19.00      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.47		0.000	10371	602288	1	
4	2.67		0.000	2180	17873	2	
7	3.05		0.000	2528	27860	2	
8	3.24		0.000	2057	24285	2	
14	4.88		0.000	2615	19356	2	
16	5.12		0.000	3388	22436	2	
17	5.28		0.000	3671	69490	2	
26	7.44		0.000	1600	21361	2	
31	9.07	formate	0.044	1893	34502	1	0.00
32	9.71		0.000	2088	43130	1	
Totals			0.044	32390	882580		

File: 99030911.D05 Sample: PREP BLK

```

=====
Sample Name: S98T002613                      Date: 03/10/1999 00:16:16
Data File  : C:\DX\DATA\99030911.D06
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 6           Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 21 5700 5Hz 0.00 19.00 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	5409	218634	1	
3	1.73		0.000	14361	143792	1	
4	2.81		0.000	4376	81959	2	
5	3.09		0.000	3580	52598	2	
14	4.71		0.000	1582	20822	2	
16	5.12		0.000	1809	19103	2	
18	5.45		0.000	3276	40612	2	
19	5.84		0.000	3843	41095	2	
20	5.96		0.000	4073	66363	2	
23	7.20	fluoride	2.931	356330	5531412	2	0.00
25	8.37	acetate	1.129	14130	270903	2	0.00
26	9.09	formate	1.294	16223	312698	2	0.00
Totals			5.354	428990	6799992		

File: 99030911.D06 Sample: S98T002613

Sample Name: S98T002613 DUP Date: 03/10/1999 00:39:16
 Data File : C:\DX\DATA\99030911.D07
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 7 Detector: CDM-2
 Analyst : Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.30		0.000	4092	32708	2	
2	0.50		0.000	1543	26204	2	
3	1.73		0.000	15434	157564	1	
5	2.93		0.000	6800	138208	2	
6	3.17		0.000	6475	50918	2	
7	3.24		0.000	6411	162229	2	
8	3.75		0.000	4552	25830	2	
9	3.88		0.000	4305	23437	2	
10	3.97		0.000	4152	24923	2	
11	4.12		0.000	3600	33608	2	
12	4.27		0.000	2249	19994	2	
13	4.51		0.000	2324	19546	2	
14	4.67		0.000	2387	24834	2	
23	6.09		0.000	2151	17738	2	
26	7.20	fluoride	2.911	354455	5485290	3	0.00
27	7.81	glycolate	0.470	2540	31905	4	0.00
28	8.35	acetate	1.129	13240	270940	2	0.00
29	9.07	formate	1.300	15479	317173	2	0.00
30	9.89		0.000	2553	42080	1	
Totals			5.810	454744	6905129		

File: 99030911.D07 Sample: S98T002613 DUP

LABCORE Completed Worklist Report for Worklist# 28648

Analyst: kjt

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: SY101 FOR @IC4G-01 RTS & BY-105

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 CCB	0	@IC4G-QC F*4	QC	1	2.40e-02	0.024	ug/mL
1 CCB	0	@IC4G-QC ACETATE2	QC	1	3.30e-02	0.033	ug/mL
1 CCB	0	@IC4G-QC FORMATE2	QC	1	4.50e-02	0.045	ug/mL
1 CCB	0	@IC4G-QC GLYCOLT1	QC	1	2.30e-02	0.023	ug/mL
2 LCS-INST	0	@IC4G-QC F*4	QC	6.10e1	5.81e+01	95.246 %	Recovery
2 LCS-INST	0	@IC4G-QC ACETATE2	QC	1.16e2	1.06e+02	91.379 %	Recovery
2 LCS-INST	0	@IC4G-QC FORMATE2	QC	1.19e2	1.15e+02	96.639 %	Recovery
2 LCS-INST	0	@IC4G-QC GLYCOLT1	QC	1.08e2	9.59e+01	88.796 %	Recovery
3 CCV	0	@IC4G-QC F*4	QC	6.29e1	5.82e+01	92.528 %	Recovery
3 CCV	0	@IC4G-QC ACETATE2	QC	1.20e2	1.18e+02	98.333 %	Recovery
3 CCV	0	@IC4G-QC FORMATE2	QC	1.20e2	1.26e+02	105.000 %	Recovery
3 CCV	0	@IC4G-QC GLYCOLT1	QC	1.03e2	9.97e+01	96.796 %	Recovery
4 SAMPLE	S99T000437	@IC4G-01 F*4-01	LIQUID	N/A	1.28e+02	128.400	ug/mL
4 SAMPLE	S99T000437	@IC4G-01 FORMATE2	LIQUID	N/A	1.202e+04	256.700	ug/mL
5 DUP	S99T000437	@IC4G-01 F*4-01	LIQUID	<1.28e2	<1.28e2		RPD
5 DUP	S99T000437	@IC4G-01 ACETATE2	LIQUID	3.96e+03	2.71e+03	37.481	RPD
5 DUP	S99T000437	@IC4G-01 FORMATE2	LIQUID	1.20e+04	1.14e+04	5.128	RPD
5 DUP	S99T000437	@IC4G-01 GLYCOLT1	LIQUID	6.48e+03	2.68e+03	82.969	RPD
6 SPK	S99T000437	@IC4G-01 F*4-01	LIQUID	6.10e1	5.80e+01	95.082 %	Recovery
6 SPK	S99T000437	@IC4G-01 ACETATE2	LIQUID	1.16e2	8.39e+01	72.328 %	Recovery
6 SPK	S99T000437	@IC4G-01 FORMATE2	LIQUID	1.19e2	1.23e+02	103.361 %	Recovery
6 SPK	S99T000437	@IC4G-01 GLYCOLT1	LIQUID	1.08e2	1.13e+01	10.463 %	Recovery
7 SAMPLE	S98T002815	@IC4G-01 F*4-01	LIQUID	N/A	2.849e+02	18.790	ug/mL
7 SAMPLE	S98T002815	@IC4G-01 ACETATE2	LIQUID	N/A	2.249e+03	35.150	ug/mL
7 SAMPLE	S98T002815	@IC4G-01 FORMATE2	LIQUID	N/A	1.755e+03	37.570	ug/mL
7 SAMPLE	S98T002815	@IC4G-01 GLYCOLT1	LIQUID	N/A	3.394e+01	33.940	ug/mL
8 DUP	S98T002815	@IC4G-01 F*4-01	LIQUID	2.85e+02	2.87e+02	0.699	RPD
8 DUP	S98T002815	@IC4G-01 ACETATE2	LIQUID	2.25e+03	2.41e+03	6.867	RPD
8 DUP	S98T002815	@IC4G-01 FORMATE2	LIQUID	1.76e+03	1.82e+03	3.352	RPD
8 DUP	S98T002815	@IC4G-01 GLYCOLT1	LIQUID	<3.39e1	<3.39e1		RPD

Final page for worklist# 28648

Analyst Signature

Date

Analyst Signature

Date

James M. Fuchs
Reviewer Signature Date 3/18/99

Units shown for QC (BLK/BKG) may not reflect the actual units.

HNF-1663 REV. 0

Page: 1

03/15/99 09:31

ws2

LABCORE Data Entry Template for Worklist# 28648 Analyst: KST Instrument: IC 4500 Book# 59W21AMethod: LA-533-105 Rev/Mod F-0 4552

Worklist Comment: SY101 FOR @IC4G-01 RTS & BY105

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	CCB				@IC4G-QC QC			
2	LCS-INST				@IC4G-QC QC			
3	CCV				@IC4G-QC QC			
4	SAMPLE	S99T000437 0			@IC4G-01 LIQUID		99000080	101SY DILUT
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
5	DUP	S99T000437 0			@IC4G-01 LIQUID			
6	SPK	S99T000437 0			@IC4G-01 LIQUID			
7	SAMPLE	S98T002815 0			@IC4G-01 LIQUID		98000437	BY-105 (2)
		Analytes Requested: ACETATE2, F*4-01, FORMATE2, GLYCOLT1						
8	DUP	S98T002815 0			@IC4G-01 LIQUID			

Final page for worklist # 28648

K. Thomdike 03-15-99
 Signature Date

_____
 Signature Date

Data Entry Comments:

uploaded 3-17-99John Howell28648MAR.C3V

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

1819

Acetate and Glycolate rejected
for poor RPD & Spk. Rest Validated
2/17/99 JM Lyle

Reanalyzed on Wt# 28648. 3/24/99

HNF-1663 REV. 0

=====

Sample Name: ICB Date: 03/15/1999 11:20:00

Data File : C:\DX\DATA\99031501.D02

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 2 Detector: CDM-2

Analyst : *KS Thordike* Column: AG14A-SC, AS14A-SC, SRS

=====

=====

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
external	1	1	5377	5Hz	0.00	17.92	16000	

=====

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	16860	552670	1	
2	1.77		0.000	67046	844872	1	
3	2.80		0.000	3197	41206	2	
14	4.28		0.000	4356	69768	1	
16	4.99		0.000	1338	20298	1	
22	6.16		0.000	4284	65796	2	
23	7.01	fluoride	0.024	2622	62409	2	0.00
24	7.41		0.000	2416	30581	2	
25	7.76	glycolate	0.023	1676	39105	2	0.00
26	8.21		0.000	1952	17241	2	
27	8.40	acetate	0.033	4783	78138	2	0.00
29	9.01	formate	0.045	2251	49293	2	0.00
Totals			0.126	112779	1871378		

File: 99031501.D02 Sample: ICB

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1820 TO 1826.

=====

Sample Name: ICV 59N21A	Date: 03/15/1999 11:37:27
Data File : C:\DX\DATA\99031501.D03	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 3	Detector: CDM-2
Analyst : <i>Ks Thorndike</i>	Column: AG14A-SC, AS14A-SC, SRS

=====

alibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
external	1	101	4976	5Hz	0.00	16.58	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	4366	70503	2	
2	0.60		0.000	2578	22863	2	
6	2.81		0.000	2289	36124	2	
8	3.13		0.000	2443	39549	2	
13	4.28		0.000	3603	43299	2	
19	5.24		0.000	1823	19198	2	
24	6.40		0.000	1494	18706	2	
25	7.20	fluoride	<i>95.3%</i> 58.135	1821383	26436634	2	0.00
26	7.87	glycolate	<i>38.7%</i> 95.864	477965	8669814	2	0.00
27	8.35	acetate	<i>91.0%</i> 105.662	490361	9648380	2	0.00
28	9.04	formate	<i>96.2%</i> 114.549	926198	17538708	2	0.00
Totals			374.210	3734503	62543777		

File: 99031501.D03 Sample: ICV 59N21A

=====

Sample Name: CCV 58N21A Date: 03/15/1999 11:55:20

Data File : C:\DX\DATA\99031501.D04

Method : C:\DX\METHOD\MAS14GLY.MET

ACI Address: 1 System: 2 Inject#: 4 Detector: CDM-2

Analyst : *Ks Thorndike* Column: AG14A-SC, AS14A-SC, SRS

=====

Calibration Volume Dilution Points Rate Start Stop Area Reject

-----

External 1 101 5120 5Hz 0.00 17.06 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.20		0.000	3393	24326	1	
2	2.24		0.000	2089	16632	1	
6	2.95		0.000	2665	60848	2	
8	3.63		0.000	3819	57332	1	
24	7.15	fluoride	<i>92.59</i> 58.197	1797802	26466210	2	0.00
25	7.84	glycolate	<i>96.87</i> 99.665	494608	9028315	2	0.00
26	8.32	acetate	<i>98.29</i> 117.816	561250	10777594	2	0.00
27	9.01	formate	<i>104.89</i> 125.822	1022356	19359087	3	0.00
28	9.71		0.000	2495	44506	4	
Totals			401.500	3890478	65834849		

File: 99031501.D04 Sample: CCV 58N21A

Sample Name: S99T000437 Date: 03/15/1999 13:36:47
 Data File : C:\DX\DATA\99031501.D05
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 5 Detector: CDM-2
 Analyst : *KJ Thorndike* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	4141	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	13028	502273	2	
2	1.73		0.000	17350	190507	2	
5	4.05		0.000	8455	708434	2	
6	4.21		0.000	6986	130375	3	
7	4.67		0.000	1309	17366	4	
8	7.87	glycolate	6481.574	437647	14619066	2	0.00
9	8.35	acetate	3954.579	395936	8791675	2	0.00
10	9.01	formate	12023.741	2357293	47062916	3	0.00
11	9.76		0.000	82257	1535774	4	
Totals			22459.894	3320259	73558386		

File: 99031501.D05 Sample: S99T000437

Sample Name: S99T000437 DUP Date: 03/15/1999 13:57:19
 Data File : C:\DX\DATA\99031501.D06
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 6 Detector: CDM-2
 Analyst : *K. Thorndike* Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	4141	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	1593	42036	1	
2	1.70		0.000	14587	161416	1	
4	2.80		0.000	7180	98730	2	
5	2.88		0.000	6718	28182	2	
6	3.15		0.000	7024	160988	2	
7	3.36		0.000	5565	34827	2	
8	3.51		0.000	4904	23582	2	
9	3.57		0.000	5006	30545	2	
10	3.72		0.000	4485	24809	2	
11	3.89		0.000	4337	62545	2	
12	4.12		0.000	2194	28218	3	
14	4.63		0.000	2139	33049	2	
15	4.95		0.000	3055	31796	2	
16	5.08		0.000	3996	40833	2	
17	5.23		0.000	4606	60253	2	
18	5.48		0.000	6101	47151	2	
19	5.61		0.000	6983	234039	2	
20	7.20	fluoride	108.609	4334	154027	2	0.00
21	7.87	glycolate	2675.986	322100	5806925	2	0.00
22	8.35	acetate	2710.431	307659	5961741	2	0.00
23	9.04	formate	11389.942	2398766	44450614	2	0.00
24	9.79		0.000	88735	1802146	2	
Totals			16884.968	3212067	59318454		

File: 99031501.D06 Sample: S99T000437 DUP

Sample Name: S99T000437 SPIKE Date: 03/15/1999 14:17:41
 Data File : C:\DX\DATA\99031501.D07
 Method : C:\DX\METHOD\MAS14GLY.MET
 ACI Address: 1 System: 2 Inject#: 7 Detector: CDM-2
 Analyst : *Ks Thorndike* Column: AG14A-SC, AS14A-SC, SRS

alibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
external	1	4141	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret. Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	5830	222702	1	
2	1.73		0.000	15090	148066	1	
4	2.88		0.000	7297	116653	2	
5	3.01		0.000	8064	93761	2	
6	3.35		0.000	7171	125262	2	
7	3.51		0.000	6543	82694	2	
8	3.72		0.000	5066	48205	2	
9	4.03		0.000	4408	92462	2	
10	4.35		0.000	2592	16359	2	
11	4.60		0.000	3223	55033	2	
12	5.09		0.000	2242	49364	2	
16	5.72		0.000	2136	100180	2	
17	7.20	fluoride	2340.563	1784842	25934900	2	0.00
18	7.89	glycolate	6938.782	896567	15704142	2	0.00
19	8.37	acetate	7336.142	851522	16430458	2	0.00
20	9.04	formate	16978.853	3565638	67902035	3	0.00
21	9.81		0.000	86771	1638068	4	
Totals			33594.340	7255004	128760343		

File: 99031501.D07 Sample: S99T000437 SPIKE

Minutes

1826

LABCORE Completed Worklist Report for Worklist# 28718

Analyst: adp

Instrument: IC45S2

Book#: 59N21A

Method: LA-533-105 Rev/Mod F-0

Worklist Comment: Rerun BY-105 @IC4G-01 for F. Try DF6. jmf

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	CCB	0	IC14G-QC F*4	QC	1	<1.10e-2		ug/mL
1	CCB	0	IC14G-QC ACETATE2	QC	1	<2.00e-2		ug/mL
1	CCB	0	IC14G-QC FORMATE2	QC	1	1.69e-01	0.169	ug/mL
1	CCB	0	IC14G-QC GLYCOLT1	QC	1	<2.00e-2		ug/mL
2	LCS-INST	0	IC14G-QC F*4	QC	6.10e1	6.24e+01	102.295 %	Recovery
2	LCS-INST	0	IC14G-QC ACETATE2	QC	1.16e2	1.12e+02	96.552 %	Recovery
2	LCS-INST	0	IC14G-QC FORMATE2	QC	1.19e2	1.23e+02	103.361 %	Recovery
2	LCS-INST	0	IC14G-QC GLYCOLT1	QC	1.08e2	1.02e+02	94.444 %	Recovery
3	CCV	0	IC14G-QC F*4	QC	6.29e1	5.95e+01	94.595 %	Recovery
3	CCV	0	IC14G-QC ACETATE2	QC	1.20e2	1.20e+02	100.000 %	Recovery
3	CCV	0	IC14G-QC FORMATE2	QC	1.20e2	1.29e+02	107.500 %	Recovery
3	CCV	0	IC14G-QC GLYCOLT1	QC	1.03e2	9.97e+01	96.796 %	Recovery
4	BLNK-PREP	0	IC14G-01 F*4-01	SOLID	1	<3.10e-2		ug/g
4	BLNK-PREP	0	IC14G-01 ACETATE2	SOLID	1	<2.00e-2		ug/g
4	BLNK-PREP	0	IC14G-01 FORMATE2	SOLID	1	<6.20e-2		ug/g
4	BLNK-PREP	0	IC14G-01 GLYCOLT1	SOLID	1	<5.60e-2		ug/g
5	SAMPLE	S98T002614	0 W	IC14G-01 F*4-01	SOLID	N/A	1.930e+02	127.100 ug/g
6	DUP	S98T002614	0 W	IC14G-01 F*4-01	SOLID	1.93e+02	3.55e+02	59.124 RPD
6	DUP	S98T002614	0 W	IC14G-01 ACETATE2	SOLID	?	2.76e+03	RPD
6	DUP	S98T002614	0 W	IC14G-01 FORMATE2	SOLID	?	3.67e+02	RPD
6	DUP	S98T002614	0 W	IC14G-01 GLYCOLT1	SOLID	?	<2.37e2	RPD
7	SPK	S98T002614	0 W	IC14G-01 F*4-01	SOLID	6.10e1	5.66e+01	92.787 % Recovery
7	SPK	S98T002614	0 W	IC14G-01 ACETATE2	SOLID	1.16e2	0.00e+00	0.000 % Recovery
7	SPK	S98T002614	0 W	IC14G-01 FORMATE2	SOLID	1.19e2	0.00e+00	0.000 % Recovery
7	SPK	S98T002614	0 W	IC14G-01 GLYCOLT1	SOLID	1.08e2	0.00e+00	0.000 % Recovery
8	SAMPLE	S98T002615	0 W	IC14G-01 F*4-01	SOLID	N/A	6.647e+01	37.240 ug/g
9	DUP	S98T002615	0 W	IC14G-01 F*4-01	SOLID	6.65e+01	6.20e+01	7.004 RPD
9	DUP	S98T002615	0 W	IC14G-01 ACETATE2	SOLID	?	<2.33e1	RPD
9	DUP	S98T002615	0 W	IC14G-01 FORMATE2	SOLID	?	3.77e+02	RPD
9	DUP	S98T002615	0 W	IC14G-01 GLYCOLT1	SOLID	?	1.84e+03	RPD

Final page for worklist# 28718

Analyst Signature	Date

Analyst Signature	Date

Reviewer Signature Janet Lyle Date 3/16/99

LABCORE Data Entry Template for Worklist# 28718

Analyst: ADP Instrument: IC 4552 Book# 59120 ³⁻¹⁴⁻⁹⁹ 59N21-A LCSMethod: LA-533-105 Rev/Mod F-0 58N21-A CCV

Worklist Comment: Rerun BY-105 @IC4G-01 for F. Try DF6. jmf

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 CCB			@IC4G-QC	QC		
2 LCS-INST			@IC4G-QC	QC		
3 CCV			@IC4G-QC	QC		
4 BLNK-PREP			@IC4G-01	SOLID		
5 SAMPLE	S98T002614	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: F*4-01						
6 DUP	S98T002614	0 W	@IC4G-01	SOLID		
7 SPK	S98T002614	0 W	@IC4G-01	SOLID		
8 SAMPLE	S98T002615	0 W	@IC4G-01	SOLID	98000436	BY-105 (2)
Analytes Requested: F*4-01						
9 DUP	S98T002615	0 W	@IC4G-01	SOLID		

Final page for worklist # 28718

Signature3-14-99
Date

Signature

Date

Data Entry Comments:

uploaded 3-15-99in Excel

28718MAR.CSV

Dup 2614 has baseline problem so
sample result is better. Validated 3/15/99jmf

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Data Reprocessed On 03/14/1999 08:38:18

=====

Sample Name: 59N21-A	LCS	Date: 03/14/1999 08:27:13
Data File :	C:\DX\DATA\99031401.D02	
Method :	C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1	System: 2	Inject#: 2
Analyst :	Column: AG14A-SC, AS14A-SC,	SRS
		Detector: CDM-2

=====

03-14-99

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	101	5700	5Hz	0.00	19.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	5340	66352	2	
2	0.60		0.000	2597	22825	2	
6	2.69		0.000	3063	31070	2	
7	2.84		0.000	2855	34917	2	
8	3.07		0.000	4108	31037	2	
9	3.23		0.000	3350	24673	2	
10	3.45		0.000	3703	33911	2	
16	4.47		0.000	1976	19381	2	
19	5.08		0.000	648	16627	2	
25	6.93	fluoride	62.417/02.32	2033266	28488445	2	0.00
26	7.55	glycolate	102.379/4.70	547485	9284604	2	0.00
27	8.00	acetate	112.263/6.78	553712	10261929	2	0.00
28	8.64	formate	123.409/03.70	1085580	18968955	2	0.00
Totals			400.468	4247682	67284726		

File: 99031401.D02 Sample: 59N21-A LCS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1829 TO 1838

Data Reprocessed On 03/14/1999 08:57:35

```

=====
Sample Name: 58N21-A   CCV                               Date: 03/14/1999 08:49:18
Data File  : C:\DX\DATA\99031401.D03
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 3                      Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          101    5700 5Hz   0.00 19.00   16000
-----

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	3034	119896	1	
4	2.19		0.000	26826	493278	2	
5	2.28		0.000	26111	277690	2	
6	2.49		0.000	21278	254883	2	
7	2.67		0.000	18437	196397	2	
8	2.88		0.000	15964	65051	2	
9	2.95		0.000	14690	73054	2	
10	3.04		0.000	12694	249963	3	
13	3.79		0.000	2058	17332	2	
15	4.16		0.000	1463	19093	2	
16	4.43		0.000	2610	18998	2	
27	6.91	fluoride	59.530	1877626	27104687	2	0.00
28	7.55	glycolate	99.734	526867	9034788	2	0.00
29	8.00	acetate	120.375	600134	11015146	2	0.00
30	8.64	formate	128.641	1139176	19815076	2	0.00
Totals			408.279	4288968	68755331		

File: 99031401.D03 Sample: 58N21-A CCV

Data Reprocessed On 03/15/1999 10:38:17

```

=====
Sample Name: BLANK CCB                      Date: 03/14/1999 08:04:53
Data File  : E:\DATA\99031401.D01
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 1
Analyst    :                               Column: AG14A-SC,AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          1    5700 5Hz    0.00 19.00    16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.40		0.000	70791	1203492	1	
2	2.23		0.000	49396	1252864	3	
5	9.12	formate	0.169	48831	2012400	1	0.00
Totals			0.169	169018	4468756		

File: 99031401.D01 Sample: BLANK CCB

```

=====
Sample Name: PREP BLANK                      Date: 03/14/1999 09:09:02
Data File  : C:\DX\DATA\99031401.D04
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 4
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
External          1          1 5700 5Hz 0.00 19.00 16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.27		0.000	16330	629492	1	
2	2.28		0.000	2413	16232	1	
7	2.92		0.000	2329	55550	2	
17	4.64		0.000	2828	18934	2	
18	4.84		0.000	2519	23498	2	
19	5.00		0.000	2711	19525	2	
21	5.33		0.000	2889	32600	2	
30	7.44		0.000	1545	17875	2	
31	7.89	glycolate	0.021	853	18404	1	0.00
34	9.09	formate	0.043	1463	23744	1	0.00
35	9.84		0.000	2275	26036	1	
Totals			0.064	38154	881891		

File: 99031401.D04 Sample: PREP BLANK

Data Reprocessed On 03/15/1999 10:50:34

Sample Name: S98T002614 SAM	Date: 03/14/1999 09:52:37
Data File : E:\DATA\99031401.D06	
Method : C:\DX\METHOD\MAS14GLY.MET	
ACI Address: 1 System: 2 Inject#: 6	Detector: CDM-2
Analyst :	Column: AG14A-SC, AS14A-SC, SRS

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	6.19	9.00	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.33		0.000	15386	901964	1	
2	2.29		0.000	4851	80120	1	
3	2.73		0.000	5041	43851	2	
4	2.83		0.000	6015	47272	2	
5	3.07		0.000	7017	72052	2	
6	3.16		0.000	5951	41970	2	
7	3.29		0.000	5085	69484	2	
8	3.59		0.000	3941	28225	2	
9	3.71		0.000	3921	25422	2	
10	3.83		0.000	3562	45664	2	
11	4.16		0.000	2938	26835	2	
12	4.32		0.000	2159	25993	2	
13	4.51		0.000	1987	18389	2	
14	4.72		0.000	2082	16945	2	
15	4.88		0.000	2003	26108	2	
19	5.61		0.000	1660	29136	2	
24	6.91	fluoride	0.989	75673	1142836	2	0.00
25	7.97	acetate	7.374	161552	3093089	2	0.00
26	8.59	formate	2.147	54495	956460	2	0.00
27	9.28		0.000	8158	129677	2	
28	9.89		0.000	2593	51243	2	
Totals			10.510	376069	6872735		

Data Reprocessed On 03/15/1999 10:47:53

```

=====
Sample Name: S98T002614  DUP                      Date: 03/14/1999 11:09:48
Data File  : E:\DATA\99031401.D07
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 7              Detector:CDM-2
Analyst    :                      Column: AG14A-SC,AS14A-SC, SRS
=====
  
```

```

-----
Calibration Volume Dilution Points Rate Start Stop Area Reject
-----
External          1          21    5700  5Hz   5.35 10.57    16000
  
```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	27677	482222	1	
3	2.29		0.000	2471	19624	2	
4	6.93	fluoride	1.762	158145	2887063	2	0.00
5	7.97	acetate	13.687	217507	5935476	2	0.00
6	8.61	formate	1.822	45319	711082	2	0.00
7	9.31		0.000	6152	93878	1	
Totals			17.271	457271	10129346		

File: 99031401.D07 Sample: S98T002614 DUP

Data Reprocessed On 03/15/1999 10:41:14

Sample Name: S98T002614	SPK	Date: 03/14/1999 11:30:01
Data File : E:\DATA\99031401.D08		
Method : C:\DX\METHOD\MAS14GLY.MET		
ACI Address: 1	System: 2	Inject#: 8
Analyst :	Column: AG14A-SC,AS14A-SC,	SRS
		Detector:CDM-2

Calibration	Volume	Dilution	Points	Rate	Start	Stop	Area	Reject
External	1	21	5700	5Hz	5.61	10.66	16000	

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.23		0.000	1186	19848	1	
3	2.24		0.000	5514	88996	1	
4	2.72		0.000	4541	50127	2	
5	2.91		0.000	4947	60482	2	
6	3.05		0.000	4817	40782	2	
7	3.31		0.000	4719	48857	2	
8	3.41		0.000	4732	55927	2	
9	3.67		0.000	3114	22789	2	
10	3.80		0.000	2724	28420	2	
11	4.11		0.000	1942	32548	2	
12	4.40		0.000	1562	18761	2	
21	5.77		0.000	1325	21644	1	
23	6.93	fluoride	12.310	1903266	26948678	2	0.00
24	7.57	glycolate	19.737	493338	8581296	2	0.00
25	8.03	acetate	28.404	683484	12520538	2	0.00
26	8.67	formate	25.273	1048355	18668733	3	0.00
27	9.33		0.000	5071	73419	4	
28	9.92		0.000	1700	20056	4	
Totals			85.724	4176336	67301900		

File: 99031401.D08 Sample: S98T002614 SPK

Data Reprocessed On 03/15/1999 10:55:54

```

=====
Sample Name: S98T002615 SAM                      Date: 03/14/1999 12:17:37
Data File  : E:\DATA\99031401.D10
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1 System: 2 Inject#: 10              Detector: CDM-2
Analyst    :                               Column: AG14A-SC, AS14A-SC, SRS
=====

```

```

=====
Calibration Volume Dilution Points Rate Start Stop Area Reject
=====
External          1           6 5700 5Hz 5.91 9.75      16000
=====

```

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
1	0.37		0.000	6107	180472	1	
3	1.73		0.000	46380	572080	1	
4	2.99		0.000	10770	248569	2	
5	3.12		0.000	9201	68766	2	
6	3.25		0.000	8072	128246	2	
7	3.56		0.000	3496	19253	2	
8	3.68		0.000	2560	20355	3	
22	5.63		0.000	1095	39280	1	
23	6.85	fluoride	0.332	96937	1531791	2	0.00
24	7.92	glycolate	8.894	718699	13808292	2	0.00
25	8.51	formate	1.710	218889	3861349	2	0.00
26	9.17		0.000	15570	270193	2	
27	9.71		0.000	11191	206614	2	

```

-----
Totals              10.936      1148967      20955259
-----

```

File: 99031401.D10 Sample: S98T002615 SAM

Data Reprocessed On 03/15/1999 10:57:29

```

=====
Sample Name: S98T002615  DUP                      Date: 03/14/1999 12:38:15
Data File  : E:\DATA\99031401.D11
Method     : C:\DX\METHOD\MAS14GLY.MET
ACI Address: 1  System: 2  Inject#: 11
Analyst    :                      Column: AG14A-SC, AS14A-SC, SRS
Detector: CDM-2
=====

```

Calibration Volume Dilution Points Rate Start Stop Area Reject

External 1 6 5700 5Hz 5.83 10.30 16000

***** Peak Report: All Peaks *****

Pk. Num	Ret Time	Component Name	Concentration ug/ml	Height	Area	Bl. Code	%Delta
4	1.73		0.000	43500	484760	1	
5	3.03		0.000	5160	191471	2	
14	4.35		0.000	3239	44293	2	
15	4.52		0.000	4170	18601	2	
16	4.71		0.000	5348	89117	2	
17	5.00		0.000	6403	184623	2	
18	5.43		0.000	6568	42369	2	
19	5.59		0.000	7413	114962	2	
20	5.83		0.000	6859	73279	2	
21	6.88	fluoride	0.319	81657	1428951	2	0.00
22	7.97	glycolate	9.451	739722	14715764	2	0.00
23	8.61	formate	1.939	230722	4469609	3	0.00
24	9.28		0.000	12177	185428	4	
25	9.84		0.000	11700	209028	4	
Totals			11.708	1164639	2252257		

File: 99031401.D11 Sample: S98T002615 DUP

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 6 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 1839 THRU 2014)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INDUCTIVELY COUPLED PLASMA ANALYSIS, ICP WORKLIST		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635
Specific Gravity Analysis (SpG)	
SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684
Hydroxide Analysis (OH)	
OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775
Ammonia Analysis (NH3)	
NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992
Chromium VI Analysis (CRVI)	
CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3396, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 7
of 17
Cont.

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Gamma Energy Analysis (GEA)	
GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Strontium 90 Analysis (Sr90)	
Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Plutonium 239 Analysis (Pu239)	
Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

08/20/98 10:02
A-0004-1

File # 980903B.7YT

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 25720

Analyst:

J. S. C.

Instrument: ICP01 9-3-98

Book#

738986

Method: LA-505-151/161 Rev/Mod C-2

Worklist Comment: ICP BY-105 LiBr and Field Blank (DIRECT)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T002391 0 D	@ICP-D01	LIQUID	98000436	BY-105 (2)
		Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01				
7	DUP	S98T002391 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T002392 0 D	@ICP-D01	LIQUID	98000436	BY-105 (2)
		Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01				
9	DUP	S98T002392 0 D	@ICP-D01	LIQUID		
10	ICSA		@ICP-QC	QC		
11	ICSAB		@ICP-QC	QC		
12	CCV		@ICP-QC	QC		

Data Entry Comments:

uploaded 9-4-98
John W. WallValidated by:
Saul R. Pang
9/9/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

08/20/98 10:02

A-0004-1

HNF-1663 REV. 0

Page: 2

LABCORE Data Entry Template for Worklist# 25720

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 CCB			@ICP-QC	QC		

Final page for worklist # 25720

Analyst Signature	Date
<i>JK [Signature]</i>	09-03-98
598700391	Direct, JF 1
598700391	Direct, 1
598700392	19, JF 10
598700392	19, 10

Analyst Signature	Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1840

HNF-1663 REV. 0

Analysis Report

Summary

Thu 09-03-98 01:34:38 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	980903B	ICP2	09/03/98	12:47	DKS	Q	CONC
2	ICB	980903B	ICP2	09/03/98	12:51	DKS	Q	CONC
3	LLS	980903B	ICP2	09/03/98	12:58	DKS	Q	CONC
4	ICSA	980903B	ICP2	09/03/98	13:02	DKS	Q	CONC
5	ICSAH	980903B	ICP2	09/03/98	13:05	DKS	Q	CONC
6	S98T002391	980903B	ICP2	09/03/98	13:08	DKS	S	CONC
7	S98T002391_D	980903B	ICP2	09/03/98	13:11	DKS	S	CONC
8	S98T002392	980903B	ICP2	09/03/98	13:14	DKS	S	CONC
9	S98T002392_D	980903B	ICP2	09/03/98	13:17	DKS	S	CONC
10	ICSA	980903B	ICP2	09/03/98	13:22	DKS	Q	CONC
11	ICSAH	980903B	ICP2	09/03/98	13:24	DKS	Q	CONC
12	CCV	980903B	ICP2	09/03/98	13:28	DKS	Q	CONC
13	CCB	980903B	ICP2	09/03/98	13:31	DKS	Q	CONC

JK

09-03-98

Worklist # 25720

B4-105

5887001391

5887001392

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1841 TO 1844

HNF-1663 REV. 0

Analysis Report

Averages

Thu 09-03-98 01:34:38 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9384	4.9105	5.1068	4.9464	4.9343	5.1311
2	ICB	-.00077	.00504	-.00692	.00189	.00004	.00028
3	LLS	.02119	.11361	.20633	.10677	.09903	.01089
4	ICSA	.00458	247.35	.03243	-.00679	.00032	.00061
5	ICSA B	.94654	245.42	.02333	-.01073	.46069	.48449
6	S98T002391	-.00093	.00872	-.00493	-.00045	.00022	.00027
7	S98T002391_D	-.00173	.00442	-.00255	.00095	.00019	.00032
8	S98T002392	.00079	.12979	.02184	.03789	.06166	.00280
9	S98T002392_D	-.02443	.03919	.01360	.00014	.05987	.00422
10	ICSA	.00272	242.68	.05418	-.01101	.00027	.00056
11	ICSA B	.94007	242.66	.05172	-.01191	.45372	.48079
12	CCV	4.9531	4.9088	5.0992	4.9317	4.9217	5.1329
13	CCB	-.00155	.00065	.00827	.00284	.00001	.00028

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0348	4.8849	5.0811	5.0176	4.9917	5.0133
2	ICB	-.01047	-.00727	-.00061	-.00164	-.00110	-.00076
3	LLS	.22236	Q.21872	.00975	.20650	.04049	.02060
4	ICSA	-.02916	249.27	.00537	.00872	-.00101	-.00330
5	ICSA B	.00486	248.59	.94575	.00393	.45922	.46802
6	S98T002391	-.01331	.23327	-.00007	.00245	-.00171	.00136
7	S98T002391_D	-.01675	.21886	-.00101	-.00431	-.00134	-.00173
8	S98T002392	-.05827	.90474	-.00674	.10575	-.01482	.00707
9	S98T002392_D	.14421	.82835	-.00208	-.06396	-.04223	-.02572
10	ICSA	-.00892	247.70	.00527	.00254	-.00087	-.00118
11	ICSA B	-.01511	247.32	.93794	.00352	.45711	.46461
12	CCV	5.0538	4.9452	5.0852	5.0236	5.0117	5.0309
13	CCB	-.00107	-.00759	-.00026	-.00231	-.00247	-.00165

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7308	.00680	4.8328	4.8431	5.0226	4.9289
2	ICB	-.00127	.00104	-.00180	.00608	.00100	.00052
3	LLS	.01684	.00081	.09834	.53449	.10383	.02097
4	ICSA	.00236	.01706	93.724	.10204	-.00277	.00253
5	ICSA B	.46214	.01889	93.040	.13241	-.00389	.98242
6	S98T002391	-.00055	.00190	-.00114	-.21817	.00066	.00136
7	S98T002391_D	-.00074	.00173	-.00150	-.08080	.00022	.00094
8	S98T002392	.03371	.01525	.00458	3.3008	.01090	1879.7
9	S98T002392_D	.01573	.00916	-.02403	-1.2800	-.01802	1887.8
10	ICSA	.00248	.01535	92.438	.16706	-.00342	.00327
11	ICSA B	.45543	.01839	92.348	.12729	-.00411	.95055
12	CCV	4.7178	.00768	4.8108	4.8093	5.0131	4.8997
13	CCB	-.00183	.00136	-.00082	-.16898	.00020	.00105

HNF-1663 REV. C

Analysis Report

Averages

Thu 09-03-98 01:34:38 PM

page 3

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.0390	4.9366	5.0021	5.1427	4.9702	5.0629
2	ICB	.00104	.00035	-.00304	-.00777	-.00079	-.00513
3	LLS	.21138	.02047	.09776	.19969	.20086	.03718
4	ICSA	253.02	-.00494	-.01763	193.80	.00242	-.00756
5	ICSAB	250.28	.44387	-.01963	192.53	.00271	.91862
6	S98T002391	.04720	.00056	-.00174	1.2345	-.00113	-.00489
7	S98T002391_D	.04777	.00047	-.00400	1.2522	-.00169	-.00108
8	S98T002392	.21240	.00602	.04061	6.1056	.02158	.18735
9	S98T002392_D	.03991	.00482	.03588	5.9838	-.05268	.10247
10	ICSA	247.94	-.00441	-.01989	187.65	-.00048	-.00745
11	ICSAB	247.32	.44110	-.02091	188.46	.00022	.91371
12	CCV	4.9790	4.9568	5.0194	5.1405	4.9572	5.0370
13	CCB	-.00240	.00051	-.00094	-.02255	-.00220	-.00036

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	4.9651	4.8972	4.9187	5.0459	4.8158	4.9296
2	ICB	.01322	-.00639	.00272	.00663	.03996	.00254
3	LLS	.42277	.19721	.20308	.11298	.23787	Q.12817
4	ICSA	.04758	.02126	-.03415	.01906	.08404	-.00278
5	ICSAB	.05387	.96730	-.03285	.00729	.08750	-.00322
6	S98T002391	.01282	-.00210	.01965	.01140	.04654	.99323
7	S98T002391_D	.01013	-.01365	.01347	.00456	.03013	.98374
8	S98T002392	.26255	-.00635	1.6463	.09100	.59572	3.1828
9	S98T002392_D	.26902	-.32587	1.6353	.06405	.50389	3.1423
10	ICSA	.06028	.02310	-.04212	.01879	.11457	.00018
11	ICSAB	.05531	.97459	-.02293	.00727	.10417	-.00208
12	CCV	5.0156	4.9391	4.9246	5.0509	4.8098	4.9380
13	CCB	.02067	-.00849	.00490	.00955	.04425	.00458

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	5.0114	4.9393	.21615	4.9936	4.8839	9.6158
2	ICB	.00949	.00008	-.00070	-.00070	-.00108	.03765
3	LLS	.20478	.02025	.01146	.02045	.38144	.49453
4	ICSA	-.01263	.00218	.00095	.00115	.00267	-.01821
5	ICSAB	-.01437	.00216	-.00012	.00114	-.01433	-.02221
6	S98T002391	.01448	.00319	-.00155	-.00070	-.01510	.05932
7	S98T002391_D	.01419	.00306	-.00237	-.00047	-.00503	.05430
8	S98T002392	.13361	.01729	.00573	.00228	-.20504	.47308
9	S98T002392_D	.08794	.01641	-.04487	-.01413	.04645	.40549
10	ICSA	-.00398	.00206	-.00449	.00160	.00057	-.01160
11	ICSAB	-.00845	.00212	-.00233	.00114	.03263	-.01275
12	CCV	4.9908	4.9349	.21887	4.9943	4.8885	9.5889
13	CCB	.01206	.00012	-.00295	-.00094	-.01328	.04563

HNF-1663 REV. 0

Analysis Report

Averages

Thu 09-03-98 01:34:38 PM

page 4

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9406	.01511	4.7171	4.9532
2	ICB	.00227	.00053	.00037	.00229
3	ILS	.10096	.00057	.01939	.02011
4	ICSA	.00282	.00745	-.00466	-.00232
5	ICSAB	.46460	.00753	.90776	-.00230
6	S98T002391	.00313	.00083	.00376	.00265
7	S98T002391_D	.00260	.00076	.00218	.00239
8	S98T002392	.03686	.00841	-1.2058	.03499
9	S98T002392_D	.02815	.00531	-1.2107	.02918
10	ICSA	.00341	.00776	-.00428	-.00118
11	ICSAB	.46169	.00760	.90652	-.00143
12	CCV	4.9527	.01559	4.7438	4.9484
13	CCB	.00262	.00068	-.00003	.00278

JK LBS
09-03-98

File #: 981211B.TXT

HNF-1663 REV. 0

10/22/98 08:53

A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26691

Analyst: JK Seto Instrument: ICP0412 #16048 Book# 77B30Method: LA-505-151/161 Rev/Mod C-3~~h-ok-77~~

Worklist Comment: ICP BY-105 (DIRECT)

Ensure dose rate of 30Cm is ≤ 50 mrem/hr
prior to performing this analysis

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T002814 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01					
7	DUP	S98T002814 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T002815 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01					
9	DUP	S98T002815 0 D	@ICP-D01	LIQUID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002819 0 D	@ICP-D01	LIQUID		

Data Entry Comments:

Uploaded 12-11-98
JH Warell

Validated by:

Saul M. Paring
12/14/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1845

HNF-1663 REV. C

Analysis Report

Summary

Fri 12-11-98 01:46:30 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981211B	ICP2	12/11/98	10:34	DKS	Q	CONC
2	ICB	981211B	ICP2	12/11/98	10:37	DKS	Q	CONC
3	LLS	981211B	ICP2	12/11/98	10:41	DKS	Q	CONC
4	ICSA	981211B	ICP2	12/11/98	10:43	DKS	Q	CONC
5	ICSAB	981211B	ICP2	12/11/98	10:46	DKS	Q	CONC
6	S98T002814	981211B	ICP2	12/11/98	10:50	DKS	S	CONC
7	S98T002814_D	981211B	ICP2	12/11/98	10:53	DKS	S	CONC
8	S98T002815	981211B	ICP2	12/11/98	10:57	DKS	S	CONC
9	S98T002815_D	981211B	ICP2	12/11/98	11:00	DKS	S	CONC
10	CCV	981211B	ICP2	12/11/98	12:33	DKS	Q	CONC
11	CCB	981211B	ICP2	12/11/98	12:36	DKS	Q	CONC
12	S98T002819_L	981211B	ICP2	12/11/98	12:40	DKS	S	CONC
13	S98T002819	981211B	ICP2	12/11/98	12:43	DKS	S	CONC
14	S98T002819_D	981211B	ICP2	12/11/98	12:46	DKS	S	CONC
15	S98T002819_A	981211B	ICP2	12/11/98	12:50	DKS	S	CONC
16	S98T002819_X	981211B	ICP2	12/11/98	12:55	DKS	S	CONC
17	S98T002819_AX	981211B	ICP2	12/11/98	12:58	DKS	S	CONC
18	ICSA	981211B	ICP2	12/11/98	13:34	DKS	Q	CONC
19	ICSAB	981211B	ICP2	12/11/98	13:37	DKS	Q	CONC
20	CCV	981211B	ICP2	12/11/98	13:40	DKS	Q	CONC
21	CCB	981211B	ICP2	12/11/98	13:43	DKS	Q	CONC

JK

B4-105

12-11-98

598T002814

Worklist # 26691

598T002815

598T002819

Post Spike: 598T002819

$$Al = \frac{\left(\frac{105460}{6010}\right) - \left(\frac{49398}{6010}\right)}{10} \times 100 = 93.3\%$$

$$K = \frac{\left(\frac{63880}{6010}\right) - \left(\frac{5910.9}{6010}\right)}{10} \times 100 = 96.5\%$$

$$Na = \frac{\left(\frac{271870}{6010}\right) - \left(\frac{219930}{6010}\right)}{10} \times 100 = 86.4\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1847 TO 1851.

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-11-98 01:46:30 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9297	4.8640	5.0488	5.1429	4.8356	5.0789
2	ICB	.00106	.00353	-.00225	.00673	.00008	.00004
3	LLS	.01519	.10444	.18872	.10437	.09805	.01025
4	ICSA	.00189	241.92	.01689	-.00239	.00011	.00000
5	ICSAB	.94147	241.89	.03428	-.00376	.45729	.47239
6	S98T002814	14.479	20113.	-26.547	32.362	-.00436	.24812
7	S98T002814_D	14.165	20274.	-21.547	31.786	-.04542	.30286
8	S98T002815	13.260	30355.	-14.819	39.299	-.03281	.27565
9	S98T002815_D	13.993	30630.	-16.418	35.832	-.01178	.30322
10	CCV	4.9472	4.8653	5.0176	5.1338	4.8128	5.0993
11	CCB	-.00128	.00318	-.01406	.00720	-.00003	.00009
12	S98T002819_L	17.003	49629.	-53.246	36.111	-.32539	.27881
13	S98T002819	15.529	48510.	-16.691	35.543	.02359	.38582
14	S98T002819_D	15.658	48545.	-9.2745	35.832	-.02285	.41400
15	S98T002819_A	572.71	48160.	604.41	633.79	552.19	603.46
16	S98T002819_X	16.493	49398.	-112.71	72.285	-.13104	.55325
17	S98T002819_AX	57039.	105460.	59413.	58836.	58040.	58180.
18	ICSA	.00845	239.00	.04419	-.00071	.00022	.00009
19	ICSAB	.94595	241.00	.04285	.00300	.45286	.47285
20	CCV	4.9307	4.8208	4.9970	5.0991	4.7469	5.0495
21	CCB	.00130	.00288	-.01114	.00577	-.00002	.00009

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0332	5.1469	5.0829	5.0053	4.9532	4.9656
2	ICB	.01703	.00528	.00059	.01011	-.00060	.00118
3	LLS	.21977	.23268	.01018	.20136	.04022	.02013
4	ICSA	-.02555	251.10	.00415	-.00284	-.00125	-.00381
5	ICSAB	.02257	251.50	.92985	-.00348	.45567	.45908
6	S98T002814	-5.8857	17.667	1.2192	-3.4413	2.6097	1495.6
7	S98T002814_D	-2.0409	14.079	1.1849	-4.6517	2.1636	1517.2
8	S98T002815	5.8417	11.514	1.1628	-4.7903	3.5174	1527.1
9	S98T002815_D	2.5867	12.805	1.2251	-5.2143	3.2936	1536.0
10	CCV	5.0377	5.1529	5.0913	5.0135	4.9778	4.9741
11	CCB	.00219	.00108	-.00053	-.00806	.00049	-.00020
12	S98T002819_L	66.181	19.300	-1.5748	-17.193	5.5838	1617.2
13	S98T002819	-7.2143	35.458	.98963	1.0088	4.1824	1577.7
14	S98T002819_D	5.4645	16.030	.98127	-1.6336	3.0577	1567.5
15	S98T002819_A	594.60	630.41	603.73	581.30	590.21	2142.8
16	S98T002819_X	36.177	42.323	.28576	-22.087	-8.8858	1611.9
17	S98T002819_AX	59783.	60156.	58946.	58633.	58787.	60606.
18	ICSA	.00082	251.18	.00517	.00555	.00011	-.00193
19	ICSAB	.01472	253.90	.93732	.00169	.46051	.46391
20	CCV	5.0417	5.1792	5.0760	4.9407	4.9518	4.9641
21	CCB	.00883	.00342	-.00044	-.00532	.00063	.00038

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-11-98 01:46:30 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7307	.00625	4.7811	5.0340	5.0251	4.9786
2	ICB	.00050	.00081	.00168	.25345	.00115	.00042
3	LLS	.01726	.00055	.10255	Q.80374	.10121	.02151
4	ICSA	.00140	.01545	91.901	.03990	-.00522	.00201
5	ICSAB	.46205	.01796	91.831	-.00023	-.00488	.98288
6	S98T002814	-2.1549	-.25704	16.591	3999.5	-.36386	1.7856
7	S98T002814_D	-2.4354	-.13084	16.816	4047.9	-.64974	1.4662
8	S98T002815	-1.1476	.70103	17.420	4175.5	-.88865	7.5874
9	S98T002815_D	-1.3666	.48881	17.017	4144.5	-.78340	2.9322
10	CCV	4.7280	.00627	4.7965	4.9700	5.0450	4.9341
11	CCB	-.00015	-.00014	-.00007	.05892	-.00106	-.00031
12	S98T002819_L	.16193	-.11579	31.363	5163.8	-2.7863	.00212
13	S98T002819	-.05482	.38696	32.374	4848.4	-.19359	.44614
14	S98T002819_D	.41418	-.00463	32.796	4911.5	-.26691	.25548
15	S98T002819_A	549.50	1.3460	593.15	5292.1	593.14	563.80
16	S98T002819_X	-4.6273	4.0217	32.644	5910.9	-4.7627	3.1814
17	S98T002819_AX	54686.	80.936	56342.	63880.	58418.	56765.
18	ICSA	.00214	.01294	91.226	.20657	-.00363	.00286
19	ICSAB	.45639	.01351	91.973	.27995	-.00460	.95879
20	CCV	4.6528	.00767	4.7466	5.0375	4.9589	4.8912
21	CCB	-.00031	.00095	.00055	-.10240	-.00013	.00031

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1239	4.9161	5.0467	4.9225	4.8772	4.9353
2	ICB	.00399	.00035	.00231	.02769	.00110	-.00069
3	LLS	.20767	.02027	.10189	.18336	.19784	.04359
4	ICSA	248.97	-.00520	-.01379	191.66	-.00189	-.00641
5	ICSAB	248.59	.43805	-.01762	190.97	-.00243	.90012
6	S98T002814	-1.5673	.04714	31.027	200800.	-1.7420	.42389
7	S98T002814_D	-4.2061	-.05057	30.934	199680.	-2.8082	.97760
8	S98T002815	-3.7290	.32122	31.529	196420.	1.3168	.56586
9	S98T002815_D	-3.9199	.13295	31.238	199120.	-2.0460	1.3334
10	CCV	5.1544	4.9086	5.0373	4.8856	4.8947	4.9429
11	CCB	-.00611	.00003	.00262	.02223	-.00170	-.00020
12	S98T002819_L	-4.2479	-.16887	39.996	223510.	-1.9979	-3.4525
13	S98T002819	-2.4643	.25924	35.277	213110.	-1.6897	1.3995
14	S98T002819_D	-5.5999	.22680	35.371	215050.	-1.8579	-.19615
15	S98T002819_A	592.13	570.75	634.91	208570.	569.82	576.81
16	S98T002819_X	-1.9665	2.0512	29.768	219930.	-29.305	15.423
17	S98T002819_AX	60601.	58838.	58887.	271870.	56494.	58682.
18	ICSA	246.55	-.00495	-.01313	185.81	-.00200	-.00934
19	ICSAB	248.25	.43974	-.01862	186.72	.00089	.91016
20	CCV	5.0892	4.8877	5.0200	4.8150	4.7863	4.9361
21	CCB	-.00567	.00029	.00099	.02041	-.00018	.00186

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-11-98 01:46:30 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0837	5.0180	4.9089	5.0845	4.8487	5.1587
2	ICB	.00367	.00487	.00716	-.00132	.02341	.00554
3	LLS	.40905	.21055	.22966	.09398	.22599	.14937
4	ICSA	.03906	.02164	-.03180	.00297	.05587	-.00537
5	ICSAB	.03745	.98448	-.02990	-.01682	.07565	-.00346
6	S98T002814	759.68	63.521	799.61	-5.0160	45.191	200.95
7	S98T002814_D	338.78	63.721	798.22	-1.8072	38.444	186.37
8	S98T002815	406.60	80.822	838.23	-2.1816	57.701	243.22
9	S98T002815_D	408.77	69.074	845.04	-4.8322	50.100	218.57
10	CCV	5.0550	5.0264	4.8978	5.0898	4.8731	5.1724
11	CCB	.00614	.00527	.00698	-.00351	.01343	.00094
12	S98T002819_L	497.66	77.772	906.04	-22.192	84.907	253.12
13	S98T002819	509.19	77.571	888.47	-4.5519	62.988	253.44
14	S98T002819_D	496.92	80.846	882.69	-2.0778	55.538	255.78
15	S98T002819_A	1100.5	675.29	1463.6	605.58	669.56	885.54
16	S98T002819_X	537.44	55.863	920.24	-48.480	168.89	365.42
17	S98T002819_AX	60040.	60150.	59788.	59478.	58052.	61161.
18	ICSA	.03712	.02786	-.03663	.00083	.07873	-.00084
19	ICSAB	.04698	.99568	-.01319	.01428	.09071	.00074
20	CCV	5.0440	5.0494	4.9502	5.0722	4.8620	5.1592
21	CCB	.01106	.00221	.00829	-.00693	.01802	.00496

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8898	4.8890	.19542	5.0747	4.8009	9.6948
2	ICB	.01088	.00008	.00341	.00024	-.00688	.02294
3	LLS	.19967	.02011	.00463	.01988	.39205	.48879
4	ICSA	-.01870	.00186	-.00867	.00039	-.02088	.01094
5	ICSAB	-.01564	.00189	-.01119	.00108	-.00793	.00831
6	S98T002814	-2.9638	.74019	-1.6428	-.01101	-6.0877	.84409
7	S98T002814_D	-2.7683	.73973	-2.1416	-.43168	.15706	7.5125
8	S98T002815	5.7854	.87046	-4.0058	.53400	-17.975	26.318
9	S98T002815_D	6.0277	.84419	-3.8335	.12071	-14.096	23.694
10	CCV	4.9039	4.9129	.19553	5.1076	4.7493	9.7184
11	CCB	.00012	-.00008	-.00284	-.00069	-.03224	-.01373
12	S98T002819_L	-10.380	.39969	-5.8282	-.70989	-50.740	-39.450
13	S98T002819	1.9840	.60694	.00895	-.13795	-1.6423	4.5904
14	S98T002819_D	-1.4831	.52929	1.2687	.00641	-16.281	-7.3669
15	S98T002819_A	578.12	575.26	16.273	598.72	538.64	1137.2
16	S98T002819_X	51.345	1.3098	-17.733	-2.8311	-90.756	133.03
17	S98T002819_AX	56533.	57671.	2362.1	58257.	57901.	110380.
18	ICSA	-.01107	.00177	.00141	.00067	-.00379	.01513
19	ICSAB	-.01567	.00190	.00005	.00182	-.01700	-.00644
20	CCV	4.8242	4.8285	.19991	5.0470	4.7549	9.5672
21	CCB	.00919	.00008	-.00239	-.00023	-.00611	.02314

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-11-98 01:46:30 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9733	.00740	4.7088	4.8598
2	ICB	.00154	.00030	-.00026	.00034
3	LLS	.10046	.00041	.01971	.01928
4	ICSA	.00011	.00697	-.00406	-.00370
5	ICSAB	.46146	.00720	.91907	-.00331
6	S98T002814	-.68234	-.42430	.21045	-.27356
7	S98T002814_D	-.58566	-.23871	.03106	-.05121
8	S98T002815	1.0919	.17723	-1.0468	1.0056
9	S98T002815_D	.99001	.13124	-1.0925	.99968
10	CCV	4.9844	.00746	4.7272	4.8809
11	CCB	-.00103	-.00022	.00025	-.00009
12	S98T002819_L	-2.1074	-.68980	-3.6545	.45231
13	S98T002819	.47176	-.00195	-1.5089	.34971
14	S98T002819_D	-.04325	-.27949	-2.1985	-.10839
15	S98T002819_A	588.82	1.0240	567.54	580.01
16	S98T002819_X	10.437	1.8276	.21623	6.8597
17	S98T002819_AX	58266.	92.987	56005.	58100.
18	ICSA	.00195	.00743	-.00391	-.00221
19	ICSAB	.46424	.00744	.92750	-.00291
20	CCV	4.9521	.00792	4.7388	4.8221
21	CCB	.00156	.00022	.00023	.00141

12-11-98

10/22/98 08:59

A-0004-1

File #: 980176.TXT
12-17-98 Jw

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26693

Analyst: JKScto Instrument: ICP012 # 1101-98

Book# 77830

Method: LA-505-151/161 Rev/Mod C-3

Worklist Comment: ICP BY-105 (DIRECT)

Ensure dose rate at each is ≤ 50 hr/cm²/hr
prior to performing this analysis.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T002816 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01						
7	DUP	S98T002816 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T002817 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01						
9	DUP	S98T002817 0 D	@ICP-D01	LIQUID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002939 0 D	@ICP-D01	LIQUID		

Data Entry Comments:

up loaded 12-17-98
J. Libanelli

Validated by:

Sal M. Pang
12/17/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

Page: 2

10/22/98 08:59

A-0004-1

LABCORE Data Entry Template for Worklist# 26693

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 SAMPLE	S98T002939	0 D	@ICP-D01	LIQUID	98000490	BY-105 (2)
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
14 DUP	S98T002939	0 D	@ICP-D01	LIQUID		
15 SPK (1ppm)	S98T002939	0 D	@ICP-D01	LIQUID		
16 ICSA	SPK 11/10/98		@ICP-QC	QC		
17 ICSAB			@ICP-QC	QC		
18 CCV			@ICP-QC	QC		
19 CCB			@ICP-QC	QC		

Final page for worklist # 26693

11-17-98
Analyst Signature   **Date**
 5487002816 , 025-15 , DF 601
 5487001816-D , 025-15 , 601
 5487001817 , 025-15 , DF 601
 5487001817-D , 025-15 , 601
 5487001934-L , 025-15-2-8 , DF 3005
 5487001934 , 025-15 , 601
 5487001934-D , 025-15 , 601
 5487001934-Q , 025-15 , 601
 5487001934-X , 025-15-1-8 , 6010
 5487001934-QX , 025-15-1-8 , 6010

Analyst Signature   **Date**

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1853

HNF-1663 REV. ()

Analysis Report

Summary

Thu 12-17-98 10:34:38 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981217A	ICP2	12/17/98	08:58	DKS	Q	CONC
2	ICB	981217A	ICP2	12/17/98	09:01	DKS	Q	CONC
3	LLS	981217A	ICP2	12/17/98	09:03	DKS	Q	CONC
4	ICSA	981217A	ICP2	12/17/98	09:06	DKS	Q	CONC
5	ICSAB	981217A	ICP2	12/17/98	09:09	DKS	Q	CONC
6	S98T002816	981217A	ICP2	12/17/98	09:12	DKS	S	CONC
7	S98T002816_D	981217A	ICP2	12/17/98	09:15	DKS	S	CONC
8	S98T002817	981217A	ICP2	12/17/98	09:19	DKS	S	CONC
9	S98T002817_D	981217A	ICP2	12/17/98	09:22	DKS	S	CONC
10	CCV	981217A	ICP2	12/17/98	09:29	DKS	Q	CONC
11	CCB	981217A	ICP2	12/17/98	09:32	DKS	Q	CONC
12	S98T002939_L	981217A	ICP2	12/17/98	09:35	DKS	S	CONC
13	S98T002939	981217A	ICP2	12/17/98	09:39	DKS	S	CONC
14	S98T002939_D	981217A	ICP2	12/17/98	09:42	DKS	S	CONC
15	S98T002939_A	981217A	ICP2	12/17/98	09:46	DKS	S	CONC
16	S98T002939_X	981217A	ICP2	12/17/98	09:50	DKS	S	CONC
17	S98T002939_AX	981217A	ICP2	12/17/98	09:54	DKS	S	CONC
18	ICSA	981217A	ICP2	12/17/98	10:18	DKS	Q	CONC
19	ICSAB	981217A	ICP2	12/17/98	10:21	DKS	Q	CONC
20	CCV	981217A	ICP2	12/17/98	10:25	DKS	Q	CONC
21	CCB	981217A	ICP2	12/17/98	10:31	DKS	Q	CONC

JR L&D

12-17-98

Worklist # 26643

BY-105

598T002816

598T002817

598T002939

S98T002939

Post Spike:

$$Al = \frac{\left(\frac{107870}{6010}\right) - \left(\frac{49906}{6010}\right)}{10} \times 100 = 96.4\%$$

$$Na = \frac{\left(\frac{281580}{6010}\right) - \left(\frac{223080}{6010}\right)}{10} \times 100 = 97.3\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1854 TO 1858

HNF-1663 REV. 0

Analysis Report

Averages

Thu 12-17-98 10:34:38 AM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9773	4.8872	5.0318	5.1346	4.8224	5.0705
2	ICB	.00063	-.00390	-.00637	.00696	.00013	-.00004
3	LLS	.02119	.10099	.19980	.10681	.09705	.01018
4	ICSA	.00031	243.47	.02619	-.00004	.00013	.00013
5	ICSAB	.94579	241.48	.00504	-.00331	.45382	.47152
6	S98T002816	14.296	35921.	-27.005	32.649	.08493	.37099
7	S98T002816_D	12.775	36293.	-23.909	31.534	-.09699	.34563
8	S98T002817	13.326	32575.	-21.394	32.648	-.01850	.37251
9	S98T002817_D	13.772	32470.	-21.950	33.765	.04176	.39797
10	CCV	4.9814	4.9118	5.0787	5.1675	4.8799	5.1371
11	CCB	-.00165	-.00566	-.00911	.00324	-.00006	.00008
12	S98T002939_L	8.1784	49994.	-55.002	18.170	-.43775	.53538
13	S98T002939	14.069	49230.	-6.2912	22.082	.01542	.39828
14	S98T002939_D	14.187	48449.	-16.712	21.522	.05871	.42461
15	S98T002939_A	493.19	49058.	583.30	614.00	544.33	588.92
16	S98T002939_X	12.728	49906.	-23.942	25.156	-.23979	.53310
17	S98T002939_AX	57613.	107870.	60348.	59182.	60118.	59274.
18	ICSA	.00025	239.38	.05209	.01310	.00014	.00018
19	ICSAB	.94160	239.14	-.01123	.00610	.44981	.46263
20	CCV	4.9407	4.8411	4.9835	5.0371	4.8092	5.0154
21	CCB	-.00216	-.00715	-.00601	.00975	-.00010	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9974	5.1638	5.0954	4.9850	4.9551	4.9980
2	ICB	.02685	.01100	.00028	-.00003	-.00351	-.00112
3	LLS	.21820	.23630	.00994	.20002	.04070	.02122
4	ICSA	.01711	253.03	.00465	-.00827	-.00341	-.00355
5	ICSAB	.00735	254.82	.93837	-.00103	.45886	.46450
6	S98T002816	13.406	15.385	.77532	1.1908	3.6482	1643.7
7	S98T002816_D	-7.1881	13.086	.15129	-8.1577	2.7010	1654.0
8	S98T002817	13.647	13.493	.57752	-7.9760	2.6265	1747.3
9	S98T002817_D	13.944	19.945	.99158	-3.2994	2.4814	1751.5
10	CCV	5.0223	5.1687	5.1356	5.0366	4.9815	5.0111
11	CCB	.02002	.00672	-.00153	-.00843	-.00097	-.00010
12	S98T002939_L	40.906	.55068	-3.8465	-41.999	-2.5776	1056.0
13	S98T002939	21.483	16.106	.18079	-6.2486	2.4739	1029.1
14	S98T002939_D	13.103	12.944	-.06686	-3.3031	4.4434	1026.4
15	S98T002939_A	599.15	626.23	596.02	567.69	579.42	1609.5
16	S98T002939_X	-2.8801	7.0600	-4.0163	-25.522	-8.0350	1051.8
17	S98T002939_AX	60454.	59949.	59471.	60109.	59014.	60413.
18	ICSA	-.01216	253.86	.00334	-.00270	-.00110	-.00316
19	ICSAB	-.01909	253.19	.92453	-.00014	.45148	.45844
20	CCV	5.0171	5.1311	5.0558	4.9513	4.9126	4.9453
21	CCB	.01800	-.00657	-.00130	-.01103	-.00206	-.00070

HNF-1663 REV. 0

Analysis Report

Averages

Thu 12-17-98 10:34:38 AM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.6983	.00845	4.8162	4.8639	5.0070	5.0196
2	ICB	-.00002	.00008	.00045	-.20205	-.00067	-.00020
3	LLS	.01788	.00035	.09716	.40630	.10029	.02126
4	ICSA	.00255	.02511	92.552	-.22988	-.00634	.00173
5	ICSAB	.45626	.02286	92.301	-.10708	-.00483	.97012
6	S98T002816	-1.0069	.75798	23.243	4415.2	-.04278	2.8762
7	S98T002816_D	-.84053	.69607	22.707	4476.5	-.1.7333	2.8764
8	S98T002817	-1.7599	.43229	19.595	4694.3	-1.3224	.85658
9	S98T002817_D	-1.6781	.46316	20.150	4546.2	-.84567	.85628
10	CCV	4.7616	.00777	4.8357	5.0395	5.0780	5.0154
11	CCB	.00101	-.00011	-.00053	-.11548	-.00203	-.00020
12	S98T002939_L	1.3470	-1.8877	126.93	4391.1	-7.4173	-2.1374
13	S98T002939	.77396	.08377	125.23	4814.6	-1.1856	-.18339
14	S98T002939_D	1.5247	.34252	124.42	4566.4	-.61172	.18353
15	S98T002939_A	537.87	1.5655	679.77	5382.2	581.13	550.97
16	S98T002939_X	1.0161	-1.6199	129.40	3545.3	-9.2088	-2.4434
17	S98T002939_AX	56291.	89.815	57145.	65740.	60061.	58787.
18	ICSA	.00205	.01866	91.643	-.13733	-.00610	.00234
19	ICSAB	.45037	.02211	91.324	-.13241	-.00547	.95566
20	CCV	4.6742	.00802	4.7553	4.5647	4.9879	4.9006
21	CCB	-.00026	.00003	-.00120	-.37424	-.00297	.00010

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1974	4.9186	5.0231	4.9853	4.8820	4.9125
2	ICB	.00069	-.00005	.00481	.02011	-.00435	-.00376
3	LLS	.20718	.01994	.10287	.21169	.19854	.03665
4	ICSA	251.22	-.00555	-.01492	191.84	-.00388	-.01153
5	ICSAB	248.51	.44023	-.01498	188.95	-.00305	.91180
6	S98T002816	-.86304	.11195	35.370	204440.	-.88388	-.62307
7	S98T002816_D	-11.627	.07009	34.889	205620.	-2.2445	-1.5226
8	S98T002817	-9.9224	-.05247	36.619	212320.	-2.0809	-3.3569
9	S98T002817_D	-8.0024	.01902	37.490	211100.	-2.1289	-2.8789
10	CCV	5.1795	4.9394	5.0593	4.9535	4.9399	4.9808
11	CCB	-.01411	-.00011	.00320	.03000	-.00199	-.00024
12	S98T002939_L	-37.461	-.53006	38.016	223380.	-6.6649	-23.245
13	S98T002939	-8.5060	.12522	35.071	215430.	-2.7495	-2.9436
14	S98T002939_D	-6.3857	.17550	35.458	208390.	-.12282	-1.8490
15	S98T002939_A	580.22	562.56	621.81	209880.	561.30	576.52
16	S98T002939_X	-24.324	-1.5099	35.625	223080.	-15.768	-25.294
17	S98T002939_AX	61801.	59285.	59234.	281580.	58486.	59088.
18	ICSA	246.63	-.00507	-.01300	185.28	-.00379	-.01096
19	ICSAB	245.89	.43567	-.01417	185.79	-.00162	.89885
20	CCV	5.0960	4.8747	4.9924	4.8503	4.8436	4.9208
21	CCB	-.01879	-.00011	.00401	.00743	-.00432	.00353

HNF-1663 REV. 0

Analysis Report

Averages

Thu 12-17-98 10:34:38 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0572	5.0280	4.9262	5.0869	4.8293	5.2172
2	ICB	.00631	-.00435	.01377	.00544	.00950	.02891
3	LLS	.40712	.20754	.20289	.10149	.20176	Q.26393
4	ICSA	.03155	.00766	-.01701	-.00082	.03593	.00316
5	ICSAB	.03771	1.0007	-.02244	.01791	.04860	.00465
6	S98T002816	406.40	75.307	895.10	-2.4437	46.736	238.63
7	S98T002816_D	417.82	75.230	895.38	1.2306	40.542	256.52
8	S98T002817	415.87	72.713	873.99	1.7117	41.935	298.97
9	S98T002817_D	391.64	84.194	881.78	5.4571	35.701	307.20
10	CCV	5.0726	5.0514	4.9349	5.1322	4.8704	5.2107
11	CCB	.01048	.00380	.00271	.00205	.00532	.01880
12	S98T002939_L	548.66	79.228	801.50	-1.1210	53.578	373.03
13	S98T002939	543.62	87.678	776.77	3.2783	41.908	239.88
14	S98T002939_D	462.16	89.460	778.63	1.4408	48.538	202.61
15	S98T002939_A	1032.5	686.51	1360.2	602.27	633.19	831.55
16	S98T002939_X	580.94	54.153	796.06	55.502	78.116	343.50
17	S98T002939_AX	60080.	60110.	59063.	60124.	58616.	61640.
18	ICSA	.03334	.01626	-.02446	.00049	.04791	.02469
19	ICSAB	.04374	.99144	-.02481	.00790	.04036	.01197
20	CCV	4.9741	4.9871	4.8630	5.0597	4.7853	5.1155
21	CCB	.00793	-.00647	.00836	-.00421	.00253	.00683

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8650	4.8906	.21348	5.0622	4.8213	9.6095
2	ICB	-.00173	-.00004	.00130	-.00088	.01703	-.00441
3	LLS	.19576	.01979	.00492	.01989	.43147	.47755
4	ICSA	-.02159	.00181	-.00856	.00018	.01443	-.04682
5	ICSAB	-.01740	.00175	-.00039	.00089	.01080	-.05307
6	S98T002816	3.0094	.84074	-3.9369	-.15411	3.8383	18.874
7	S98T002816_D	2.4633	.84109	-9.7780	-.30943	19.993	15.822
8	S98T002817	1.0966	.91833	-5.9545	-.42721	5.7776	8.7071
9	S98T002817_D	1.0297	.91761	-6.2269	-.29727	4.7078	14.287
10	CCV	4.9373	4.9484	.22236	5.1175	4.7794	9.7314
11	CCB	-.00130	.00000	-.00710	-.00069	.02263	-.00776
12	S98T002939_L	-23.928	.90258	-21.259	-1.4051	28.045	-97.485
13	S98T002939	-3.1905	1.0984	-6.9835	-.56328	6.0802	-7.0119
14	S98T002939_D	-1.1229	.97000	-5.5644	-.02392	2.9273	2.2653
15	S98T002939_A	565.11	564.01	20.983	582.64	551.68	1102.0
16	S98T002939_X	-26.151	.26396	-15.763	-5.4040	24.885	-99.688
17	S98T002939_AX	58313.	59286.	2701.5	59257.	59031.	113360.
18	ICSA	-.01516	.00170	-.00250	.00000	-.01093	-.04609
19	ICSAB	-.01588	.00172	.00066	.00133	-.01949	-.03752
20	CCV	4.8329	4.8584	.22067	5.0188	4.7595	9.5329
21	CCB	-.00162	.00000	-.00947	-.00092	-.00657	-.00440

HNF-1663 REV. 0

Analysis Report

Averages

Thu 12-17-98 10:34:38 AM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9677	.00806	4.7440	4.8259
2	ICB	-.00034	-.00023	.00031	-.00008
3	LLS	.09889	.00026	.01981	.01856
4	ICSA	-.00006	.00730	-.00381	-.00361
5	ICSAB	.46005	.00746	.92256	-.00312
6	S98T002816	.95856	.03908	-1.1538	1.1652
7	S98T002816_D	1.0843	.13458	-1.6749	1.6050
8	S98T002817	.56858	.09050	1.2357	.96113
9	S98T002817_D	.66083	.04103	1.1338	.97537
10	CCV	4.9943	.00762	4.7247	4.8832
11	CCB	-.00065	-.00038	.00007	.00023
12	S98T002939_L	-5.5319	-1.8427	-3.8085	-2.9138
13	S98T002939	-.66258	-.28055	-1.7012	.31953
14	S98T002939_D	.14387	-.04806	-1.3674	.56686
15	S98T002939_A	577.10	1.0861	563.29	559.42
16	S98T002939_X	-6.0539	-2.3128	-5.1620	-2.3914
17	S98T002939_AX	58699.	91.719	55554.	58961.
18	ICSA	.00137	.00777	-.00363	-.00311
19	ICSAB	.45403	.00746	.91696	-.00314
20	CCV	4.9233	.00754	4.6828	4.8039
21	CCB	-.00048	-.00031	.00006	.00101

JK L&D

12-17-98

File #: 981216G.TXT

HNF-1663 REV. 0

10/22/98 09:03

A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26694

Analyst: DK S470 Instrument: ICP01 11-61-98Book# 77B30Method: LA-505-151/161 Rev/Mod C3Ensure dose rate at 30cm is ≤ 50 mrem/hr
prior to performing this analysis.Worklist Comment: 11-61-98
ICP BY-105 (DIRECT)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T002818 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
7	DUP	S98T002818 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T002820 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
9	DUP	S98T002820 0 D	@ICP-D01	LIQUID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002960 0 D	@ICP-D01	LIQUID		

Data Entry Comments:

uploaded 12-16-98John Wauell26694DEC.CSV

Validated by:

Saul M. Pung12/17/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

10/22/98 09:03
A-0004-1

HNF-1663 REV. 0

Page: 2

LABCORE Data Entry Template for Worklist# 26694

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 SAMPLE	S98T002960 0 D	@ICP-D01	LIQUID	98000509	BY-105	(2)
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
14 DUP	S98T002960 0 D	@ICP-D01	LIQUID			
15 SPK (1ppm) S98T002960 0 D	S98T002960 0 D	@ICP-D01	LIQUID			
16 ICSA		@ICP-QC	QC			
17 ICSAB		@ICP-QC	QC			
18 CCV		@ICP-QC	QC			
19 CCB		@ICP-QC	QC			

Final page for worklist # 26694

JK *12-16-98*

Analyst Signature	Date
S987001818	.015-15 JF 601
S987001818-D	.015-15 601
S987001820	.015-15 JF 601
S987001820-D	.015-15 601
S987001860-L	.015-15-1-8 JF 3005
S987001860	.015-15 601
S987001860-D	.015-15 601
S987001860-E	.015-15 601
S987001860-X	.015-15-1-9 6010
S987001860-QX	.015-15-1-9 6010

Analyst Signature	Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1860

HNF-1663 REV. 0

Analysis Report

Summary

Wed 12-16-98 11:31:31 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981216A	ICP2	12/16/98	09:12	DKS	Q	CONC
2	ICB	981216A	ICP2	12/16/98	09:16	DKS	Q	CONC
3	LLS	981216A	ICP2	12/16/98	09:19	DKS	Q	CONC
4	ICSA	981216A	ICP2	12/16/98	09:22	DKS	Q	CONC
5	ICSAB	981216A	ICP2	12/16/98	09:25	DKS	Q	CONC
6	S98T002818	981216A	ICP2	12/16/98	10:15	DKS	S	CONC
7	S98T002818_D	981216A	ICP2	12/16/98	10:19	DKS	S	CONC
8	S98T002820	981216A	ICP2	12/16/98	10:24	DKS	S	CONC
9	S98T002820_D	981216A	ICP2	12/16/98	10:28	DKS	S	CONC
10	CCV	981216A	ICP2	12/16/98	10:35	DKS	Q	CONC
11	CCV	981216A	ICP2	12/16/98	10:39	DKS	Q	CONC
12	S98T002960_L	981216A	ICP2	12/16/98	10:42	DKS	S	CONC
13	S98T002960	981216A	ICP2	12/16/98	10:45	DKS	S	CONC
14	S98T002960_D	981216A	ICP2	12/16/98	10:48	DKS	S	CONC
15	S98T002960_A	981216A	ICP2	12/16/98	10:54	DKS	S	CONC
16	S98T002960_X	981216A	ICP2	12/16/98	11:01	DKS	S	CONC
17	S98T002960_AX	981216A	ICP2	12/16/98	11:04	DKS	S	CONC
18	ICSA	981216A	ICP2	12/16/98	11:18	DKS	Q	CONC
19	ICSAB	981216A	ICP2	12/16/98	11:21	DKS	Q	CONC
20	CCV	981216A	ICP2	12/16/98	11:24	DKS	Q	CONC
21	CCB	981216A	ICP2	12/16/98	11:28	DKS	Q	CONC

JK 240

BY-105

12-16-98

Work List # 26694

598T002818

598T002820

598T002960

598T002960

Post Spike:

$$Al = \frac{\left(\frac{85327}{6010}\right) - \left(\frac{27535}{6010}\right)}{10} \times 100 = 96.2\%$$

$$Na = \frac{\left(\frac{276940}{6010}\right) - \left(\frac{217930}{6010}\right)}{10} \times 100 = 98.2\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1861 TO 1865.

HNF-1663 REV. 0

Analysis Report

Averages

Wed 12-16-98 11:31:31 AM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9623	4.8893	5.0207	5.0634	4.8368	5.0998
2	ICB	.00013	.00213	.00357	.00046	.00007	-.00039
3	LLS	.02014	.09602	.20165	.09865	.09884	.01039
4	ICSA	.00139	243.78	.00725	-.00734	.00010	.00000
5	ICSAB	.94543	242.60	.00364	-.00555	.45894	.47568
6	S98T002818	15.471	44891.	-23.308	28.673	.01910	.34355
7	S98T002818_D	15.786	45022.	-18.496	31.152	.05591	.34325
8	S98T002820	16.179	49612.	-18.163	29.778	.07516	.31731
9	S98T002820_D	15.889	49486.	-24.868	31.984	.10142	.34333
10	CCV	4.9853	4.9412	5.0949	5.1308	4.9077	5.1910
11	CCV	.00061	.00338	-.00058	-.00229	.00016	-.00013
12	S98T002960_L	17.308	27357.	-3.5106	28.954	.19604	.13236
13	S98T002960	15.179	26723.	-22.536	21.236	.14619	.29006
14	S98T002960_D	14.948	26695.	-17.041	23.720	.14832	.21043
15	S98T002960_A	567.69	26647.	586.84	611.79	549.80	600.25
16	S98T002960_X	12.531	27535.	-12.902	13.852	.26065	-.52334
17	S98T002960_AX	58224.	85327.	60166.	59888.	59258.	59352.
18	ICSA	.00324	243.63	.02343	-.00127	.00034	.00000
19	ICSAB	.94908	243.71	.04993	-.01140	.46140	.47892
20	CCV	4.9769	4.8984	5.1067	5.1073	4.8865	5.1610
21	CCB	-.00012	.00414	.00444	-.00137	.00006	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0073	5.1535	5.1116	4.9921	4.9559	4.9866
2	ICB	.02276	.01371	-.00021	.00289	-.00153	-.00082
3	LLS	.19995	.23120	.00870	.19729	.03816	.01909
4	ICSA	-.00140	252.64	.00301	-.00328	-.00048	-.00453
5	ICSAB	.00563	253.97	.93639	.00228	.45576	.46282
6	S98T002818	4.6296	12.324	.28069	-3.4772	3.3385	1690.7
7	S98T002818_D	2.2663	10.680	.83768	-.20483	3.5512	1694.7
8	S98T002820	10.012	8.8420	-.00604	-.98020	2.6261	1581.4
9	S98T002820_D	5.6705	8.1658	-.12767	-.57236	3.2000	1577.7
10	CCV	5.0545	5.1441	5.1233	5.0813	4.9862	5.0125
11	CCV	.01192	.01627	-.00027	.00224	-.00059	.00026
12	S98T002960_L	89.772	32.500	1.1652	7.7121	.71100	1297.2
13	S98T002960	7.1006	82.171	.59860	.03385	3.1345	1271.1
14	S98T002960_D	11.513	20.495	1.1073	1.9762	2.9263	1275.5
15	S98T002960_A	593.22	636.94	604.75	584.98	584.59	1829.4
16	S98T002960_X	90.023	14.548	-6.5167	-32.696	-11.988	1295.4
17	S98T002960_AX	59861.	60131.	59390.	59495.	59728.	60757.
18	ICSA	-.01203	255.83	.00336	.00560	-.00225	-.00399
19	ICSAB	.00107	255.05	.94402	-.00152	.45823	.46742
20	CCV	5.0613	5.1981	5.1477	5.0334	5.0004	5.0247
21	CCB	.00447	.01373	-.00027	-.00387	-.00212	-.00115

HNF-1663 REV. 0

Analysis Report

Averages

Wed 12-16-98 11:31:31 AM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7045	.00841	4.8207	4.8674	5.0180	4.9889
2	ICB	-.00014	.00013	.00180	.01954	.00055	.00021
3	LLS	.01692	-.00044	.09945	Q.29103	.10076	.02089
4	ICSA	.00180	.03185	92.816	-.04563	-.00540	.00232
5	ICSAB	.45808	.03113	92.506	-.04179	-.00456	.99535
6	S98T002818	-.17132	.15284	30.343	4953.2	-.74620	1.1426
7	S98T002818_D	-.49244	-.18069	30.655	4968.9	-.37900	1.0795
8	S98T002820	.02744	-.10186	36.428	5048.4	-.17400	.57154
9	S98T002820_D	-.04904	.37813	36.089	4962.7	-.66812	.76135
10	CCV	4.7983	.00749	4.8500	4.8282	5.1187	5.1066
11	CCV	.00036	-.00007	.00148	.09353	.00069	.00042
12	S98T002960_L	.78302	-.33124	32.774	4255.7	-.69059	1.5872
13	S98T002960	-.95057	.53204	32.194	4154.2	.08521	1.0144
14	S98T002960_D	-1.0175	.63325	31.651	4135.7	.19749	1.3946
15	S98T002960_A	543.20	1.3833	594.03	4628.6	587.01	566.52
16	S98T002960_X	-1.6897	2.9909	27.170	3829.5	-6.5527	5.7050
17	S98T002960_AX	56222.	95.577	56874.	63850.	59319.	58820.
18	ICSA	.00201	.02706	93.085	.01900	-.00373	.00317
19	ICSAB	.46143	.03167	92.851	-.08364	-.00524	1.0038
20	CCV	4.7403	.00862	4.8270	4.9233	5.0665	5.0202
21	CCB	-.00068	.00001	.00061	-.10691	-.00014	.00042

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1869	4.9265	5.0272	4.9609	4.8894	4.9608
2	ICB	.00661	-.00003	.00235	.01542	-.00186	-.00066
3	LLS	.20398	.02001	.10299	.20670	.19649	.03948
4	ICSA	251.25	-.00560	-.01221	194.51	-.00316	-.01428
5	ICSAB	249.45	.43965	-.01446	192.91	-.00142	.91212
6	S98T002818	-5.4505	.24004	35.581	217920.	-.81208	2.1426
7	S98T002818_D	-4.3221	.25680	36.331	217140.	-1.1001	-.59096
8	S98T002820	1.0668	.00857	36.048	220560.	-1.1028	-1.9236
9	S98T002820_D	.74487	.12802	37.090	217410.	-1.3004	-.65904
10	CCV	5.2318	4.9365	5.0647	5.0257	5.0026	4.9751
11	CCV	.00758	-.00003	.00360	.03371	-.00104	.00123
12	S98T002960_L	10.577	-.22776	35.773	217520.	-1.2028	-5.9854
13	S98T002960	2.2343	.12725	30.982	207390.	-.70139	1.1297
14	S98T002960_D	5.1175	.13854	30.708	206250.	-.39428	-.27994
15	S98T002960_A	597.79	565.95	624.99	202910.	568.06	578.73
16	S98T002960_X	-9.7123	.47706	28.387	217930.	-21.386	-26.070
17	S98T002960_AX	61378.	59145.	59898.	276940.	57780.	59486.
18	ICSA	250.77	-.00534	-.01706	192.26	-.00385	-.00905
19	ICSAB	250.10	.44135	-.01509	193.54	-.00172	.91910
20	CCV	5.1678	4.9478	5.0665	4.9558	4.9299	4.9787
21	CCB	.00226	-.00003	.00282	.02223	-.00361	-.00410

HNF-1663 REV. 0

Analysis Report

Averages

Wed 12-16-98 11:31:31 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.1177	5.0490	4.8676	5.0876	4.8453	5.1856
2	ICB	.00017	-.01941	.00943	-.00333	-.00699	.00398
3	LLS	.40542	.18768	.20157	.09478	.18981	Q.24377
4	ICSA	.01776	.01050	-.03510	.01227	.03706	-.00595
5	ICSAB	.03715	.97235	-.02573	.00436	.04343	-.00169
6	S98T002818	459.34	67.946	923.74	-2.5974	46.139	263.52
7	S98T002818_D	457.47	69.074	929.92	-6.3448	48.046	261.66
8	S98T002820	532.28	70.805	938.43	2.4184	49.713	223.36
9	S98T002820_D	580.47	70.870	971.78	2.4623	56.325	225.80
10	CCV	5.1276	5.0426	4.8990	5.1304	4.9081	5.2130
11	CCV	.01013	-.01062	.00717	-.00589	.01039	.00767
12	S98T002960_L	353.19	41.753	762.42	-12.232	37.035	133.99
13	S98T002960	341.54	61.064	747.08	2.8470	46.462	127.08
14	S98T002960_D	346.70	72.643	746.57	-2.4168	49.041	127.97
15	S98T002960_A	942.91	669.01	1306.4	613.64	676.26	752.13
16	S98T002960_X	390.54	-76.144	778.07	-35.280	119.68	206.22
17	S98T002960_AX	60789.	60171.	59596.	60580.	58322.	60245.
18	ICSA	.04682	-.00124	-.02472	.01489	.06910	.00538
19	ICSAB	.04269	.97495	-.01076	.00437	.05703	-.00259
20	CCV	5.0759	5.0478	4.8666	5.1424	4.9380	5.1965
21	CCB	.00408	-.01501	.00613	-.00745	.01422	.00398

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8894	4.8989	.22722	5.0477	4.7826	9.6246
2	ICB	.00317	.00000	.00109	-.00043	-.00379	.00585
3	LLS	.19513	.02009	.00359	.01936	.39758	.46849
4	ICSA	-.02073	.00190	.00407	.00086	-.02773	-.00764
5	ICSAB	-.01533	.00190	.01126	.00089	-.02601	-.00099
6	S98T002818	-.59921	.76176	-.47005	.51947	-10.985	10.447
7	S98T002818_D	-2.7541	.71152	.35024	.12717	-15.298	3.8911
8	S98T002820	-3.2533	.71186	-.41031	.12512	-19.005	8.1334
9	S98T002820_D	.99291	.76158	-1.8195	.38291	-17.586	14.166
10	CCV	4.9824	5.0051	.23241	5.1436	4.8497	9.8266
11	CCV	-.00182	-.00004	.00216	.00000	-.01913	.00600
12	S98T002960_L	-2.6243	.76111	8.3843	.68231	-39.181	9.7938
13	S98T002960	4.2437	.85910	.94960	.52052	-3.6449	32.795
14	S98T002960_D	5.6901	.85990	2.6981	.38951	.84781	48.229
15	S98T002960_A	569.19	570.43	29.312	590.63	552.33	1146.9
16	S98T002960_X	41.208	1.5079	-25.170	-1.4573	-39.208	202.22
17	S98T002960_AX	57634.	58830.	2798.6	59614.	58352.	113500.
18	ICSA	-.01469	.00183	.01805	.00070	-.01118	-.00101
19	ICSAB	-.01845	.00186	.01227	.00134	-.03424	-.01241
20	CCV	4.9397	4.9426	.23725	5.1008	4.8254	9.7250
21	CCB	.00145	.00004	.00171	.00000	-.01364	.00716

HNF-1663 REV. 0

Analysis Report

Averages

Wed 12-16-98 11:31:31 AM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9660	.00813	4.7269	4.8426
2	ICB	.00048	.00023	.00017	.00051
3	LLS	.09828	-.00012	.01961	.01844
4	ICSA	-.00007	.00729	-.00380	-.00467
5	ICSAB	.46147	.00768	.92230	-.00413
6	S98T002818	.33570	-.00443	.18846	.36545
7	S98T002818_D	.03958	-.33137	.10634	.31259
8	S98T002820	.14732	-.14165	-1.5909	.19087
9	S98T002820_D	.93834	-.05312	-1.3807	.82191
10	CCV	5.0143	.00765	4.7246	4.9144
11	CCV	.00014	.00000	.00037	.00013
12	S98T002960_L	-.09337	.23135	.04023	.25410
13	S98T002960	1.4127	.17447	.65494	1.1425
14	S98T002960_D	1.5855	.35633	.29179	1.1805
15	S98T002960_A	583.04	1.0670	565.11	571.90
16	S98T002960_X	10.834	2.2693	-1.1163	10.720
17	S98T002960_AX	59245.	96.840	56461.	59153.
18	ICSA	.00131	.00775	-.00355	-.00427
19	ICSAB	.46373	.00768	.92367	-.00453
20	CCV	5.0013	.00825	4.7432	4.8841
21	CCB	.00031	.00000	.00050	.00016

JH

12-16-98

10/22/98 09:08

A-0004-1

File #: 981204B-7XT

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26695 Analyst: JK566 Instrument: ICP-OES 1102-98 Book# 77B30Method: LA-505-451/161 Rev/Mod C-3Worklist Comment: ICP BY-105 (DIRECT)*Ensure dose rate at 30cm is ≤ 50 uR/hr
prior to performing this analysis.*

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T002821 0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)	
Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01						
7	DUP	S98T002821 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T002961 0 D	@ICP-D01	LIQUID	98000509 BY-105 (2)	
Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01						
9	DUP	S98T002961 0 D	@ICP-D01	LIQUID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002962 0 D	@ICP-D01	LIQUID		

Data Entry Comments:

*uploaded 12-4-98
John Warell**Validated by:
Saul N. Pang
12/7/98*

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

10/22/98 09:08

A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 26695

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 SAMPLE	S98T002962 0 D	@ICP-D01	LIQUID		98000509 BY-105 (2)	
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
14 DUP	S98T002962 0 D	@ICP-D01	LIQUID			
15 SPK (1ppm)	S98T002962 0 D	@ICP-D01	LIQUID			
16 ICSA		@ICP-QC	QC			
17 ICSAB		@ICP-QC	QC			
18 CCV		@ICP-QC	QC			
19 CCB		@ICP-QC	QC			

Final page for worklist # 26695

12-04-98
 Analyst Signature Date
 S987002821 , 025-15 , JF 601
 S987002821-D , 025-15 , 601
 S987002961 , 025-15 , JF 601
 S987002961-D , 025-15 , 601
 S987002962-L , 025-15-28 , JF 3005
 S987002962 , 025-15 , 601
 S987002962-D , 025-15 , 601
 S987002962-Q , 025-15 , 601
 S987002962-X , 025-15-19 , 6010
 S987002962-GX , 025-15-19 , 6010

Analyst Signature Date
 _____

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1867

HNF-1663 REV. 0

Analysis Report

Summary

Fri 12-04-98 02:33:53 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981204B	ICP2	12/04/98	12:56	DKS	Q	CONC
2	ICB	981204B	ICP2	12/04/98	12:59	DKS	Q	CONC
3	LLS	981204B	ICP2	12/04/98	13:02	DKS	Q	CONC
4	ICSA	981204B	ICP2	12/04/98	13:05	DKS	Q	CONC
5	ICSAB	981204B	ICP2	12/04/98	13:08	DKS	Q	CONC
6	S98T002821	981204B	ICP2	12/04/98	13:11	DKS	S	CONC
7	S98T002821_D	981204B	ICP2	12/04/98	13:14	DKS	S	CONC
8	S98T002961	981204B	ICP2	12/04/98	13:18	DKS	S	CONC
9	S98T002961_D	981204B	ICP2	12/04/98	13:21	DKS	S	CONC
10	CCV	981204B	ICP2	12/04/98	13:24	DKS	Q	CONC
11	CCB	981204B	ICP2	12/04/98	13:27	DKS	Q	CONC
12	S98T002962_L	981204B	ICP2	12/04/98	13:31	DKS	S	CONC
13	S98T002962	981204B	ICP2	12/04/98	13:34	DKS	S	CONC
14	S98T002962_D	981204B	ICP2	12/04/98	13:37	DKS	S	CONC
15	S98T002962_A	981204B	ICP2	12/04/98	13:40	DKS	S	CONC
16	S98T002962_X	981204B	ICP2	12/04/98	13:47	DKS	S	CONC
17	S98T002962_AX	981204B	ICP2	12/04/98	13:51	DKS	S	CONC
18	ICSA	981204B	ICP2	12/04/98	14:21	DKS	Q	CONC
19	ICSAB	981204B	ICP2	12/04/98	14:24	DKS	Q	CONC
20	CCV	981204B	ICP2	12/04/98	14:27	DKS	Q	CONC
21	CCB	981204B	ICP2	12/04/98	14:30	DKS	Q	CONC

Handwritten signature

12.04.98

Worklist # 26695

B.Y.105

S98T002821

S98T002961

S98T002962

S98T002962 SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1868 TO 1872.

$$Cr = 1.23e + 3 \frac{ug}{ml}$$

$$Spk: Cr = \frac{\left(\frac{1799.6}{601}\right) - \left(\frac{1226.9}{601}\right)}{1} \times 100 = 95.3\%$$

Post Spike:

$$Me = \frac{\left(\frac{273330}{6010}\right) - \left(\frac{215440}{6010}\right)}{10} \times 100 = 96.3\%$$

$$Re = \frac{\left(\frac{96118}{6010}\right) - \left(\frac{38932}{6010}\right)}{10} \times 100 = 95.2\%$$

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-04-98 02:33:53 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9108	4.8673	4.9638	5.1093	4.8250	5.0789
2	ICB	.00012	.00709	.00438	-.00093	.00006	.00008
3	LLS	.02116	.10283	.19808	.09899	.09736	.01045
4	ICSA	.00233	242.71	.03579	-.01547	.00014	.00026
5	ICSAB	.94715	241.82	-.00078	-.01270	.45740	.47289
6	S98T002821	16.160	46560.	-3.4907	33.110	.07552	.42832
7	S98T002821_D	15.828	46308.	-10.220	33.109	.03919	.42816
8	S98T002961	15.795	29925.	-18.148	21.327	.14020	.48042
9	S98T002961_D	14.424	29200.	-9.5681	19.363	.07880	.40075
10	CCV	4.9699	4.8704	4.9942	5.1199	4.8409	5.0836
11	CCB	-.00011	-.00207	.01158	-.00279	.00000	-.00004
12	S98T002962_L	11.664	39061.	-44.381	26.654	-.32501	.80419
13	S98T002962	14.823	38425.	-20.441	29.461	.00012	.53453
14	S98T002962_D	13.955	36499.	-16.096	28.901	-.00129	.42759
15	S98T002962_A	541.32	38705.	578.93	627.67	555.17	597.84
16	S98T002962_X	17.658	38932.	25.450	5.6351	-.55518	1.6023
17	S98T002962_AX	55013.	96118.	59292.	59059.	58858.	58446.
18	ICSA	.00748	239.56	-.00593	-.00627	.00018	.00035
19	ICSAB	.94699	240.58	.05201	-.00709	.45640	.47032
20	CCV	4.9332	4.8332	4.9347	5.0835	4.7906	5.0199
21	CCB	.00127	-.00640	.00813	.00280	-.00009	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9832	5.0650	5.0341	4.9971	4.9069	4.9425
2	ICB	.05270	.00000	.00028	-.00097	.00146	.00008
3	LLS	.23446	.22435	.01049	.19881	.04130	.01871
4	ICSA	.01465	249.82	.00455	-.00494	.00046	-.00257
5	ICSAB	-.00149	249.92	.93140	.00411	.45316	.45888
6	S98T002821	37.687	6.9148	.31156	-1.1734	4.8583	1137.2
7	S98T002821_D	30.364	1.4340	.31809	-1.3539	5.3023	1136.0
8	S98T002961	21.404	15.492	1.8264	4.3068	4.4958	1241.8
9	S98T002961_D	13.055	7.8887	1.0493	.77672	3.3168	1204.1
10	CCV	5.0572	5.1607	5.0855	5.0052	4.9700	4.9985
11	CCB	.01960	-.00012	.00121	.00036	.00037	.00037
12	S98T002962_L	1.8159	-13.024	.06847	-27.429	2.5601	1239.9
13	S98T002962	23.983	15.120	.68849	.01224	4.8653	1226.9
14	S98T002962_D	29.267	8.0847	.35711	-2.9142	4.1259	1162.4
15	S98T002962_A	591.84	621.52	597.91	582.71	582.77	1799.6
16	S98T002962_X	130.57	-22.492	-.50958	-27.068	8.1562	1248.8
17	S98T002962_AX	60012.	60316.	58956.	59110.	58711.	60708.
18	ICSA	.01029	251.71	.00459	.00719	.00292	-.00113
19	ICSAB	.00794	251.64	.93011	.00230	.45452	.46021
20	CCV	5.0068	5.1280	5.0371	4.9463	4.9187	4.9527
21	CCB	.00741	-.00069	.00051	-.00779	.00000	-.00078

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-04-98 02:33:53 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7326	.00713	4.7766	4.7935	5.0417	5.0770
2	ICB	.00014	-.00012	.00163	-.26500	.00058	.00010
3	LLS	.01695	.00017	.09685	Q.33192	.10117	.02153
4	ICSA	.00342	.02422	91.661	-.20849	-.00522	.00184
5	ICSAB	.46408	.02610	91.366	-.04236	-.00376	1.0012
6	S98T002821	-.16312	.14022	50.984	4646.9	-.10783	-.18416
7	S98T002821_D	.03146	.12347	50.727	4417.2	-.11066	.12284
8	S98T002961	.18916	.10595	34.720	4231.7	1.1497	.36752
9	S98T002961_D	-.36617	-.08441	32.651	3971.5	.22786	.30649
10	CCV	4.7098	.00782	4.8049	4.7502	5.0493	4.9921
11	CCB	-.00009	.00042	.00112	-.02670	.00024	.00010
12	S98T002962_L	-2.3718	-1.2793	39.440	3589.9	-4.3826	-.91986
13	S98T002962	.77429	.26333	38.527	4513.2	.04181	.55204
14	S98T002962_D	.73554	.10576	36.772	4209.9	-.21270	.18367
15	S98T002962_A	549.88	1.0751	596.76	5020.1	593.53	570.16
16	S98T002962_X	1.7893	3.3035	36.808	4004.7	-2.0725	1.8346
17	S98T002962_AX	54989.	92.950	56725.	63438.	58890.	58842.
18	ICSA	.00353	.01716	91.275	.08354	-.00372	.00296
19	ICSAB	.45961	.02178	91.388	-.18852	-.00395	.98691
20	CCV	4.6605	.00840	4.7466	4.8730	4.9699	5.0104
21	CCB	-.00059	.00040	.00143	-.21266	-.00059	.00030

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1371	4.8682	5.0081	5.0415	4.9110	4.9260
2	ICB	.00160	.00012	.00144	-.00179	.00113	.00392
3	LLS	.20521	.01998	.10079	.20414	.19733	.04454
4	ICSA	249.18	-.00532	-.01763	195.12	-.00208	-.00484
5	ICSAB	247.67	.43744	-.01720	193.53	-.00364	.91440
6	S98T002821	-3.8075	.07595	30.591	220040.	.87859	3.3332
7	S98T002821_D	-5.3017	.00957	31.467	216370.	.76488	4.8478
8	S98T002961	1.9496	.06892	31.278	204030.	1.2056	2.6268
9	S98T002961_D	-2.1379	.05383	30.309	199920.	-1.5381	1.2475
10	CCV	5.1420	4.9255	5.0562	4.9427	4.9051	4.9808
11	CCB	-.00093	.00020	-.00241	-.00106	.00138	.00415
12	S98T002962_L	-23.668	-.47312	31.572	219200.	-11.079	4.1715
13	S98T002962	.04869	.04909	34.579	207640.	.39968	-.42032
14	S98T002962_D	-2.9765	.02367	30.888	197680.	.24322	3.5340
15	S98T002962_A	592.05	566.21	630.27	207020.	576.14	588.74
16	S98T002962_X	-27.057	1.1793	22.408	215440.	2.5177	17.986
17	S98T002962_AX	60969.	59099.	59002.	273330.	57308.	59492.
18	ICSA	246.06	-.00495	-.01569	189.69	.00075	-.00749
19	ICSAB	246.72	.43725	-.01833	190.94	-.00228	.91064
20	CCV	5.0898	4.8714	4.9986	4.9606	4.8338	4.9176
21	CCB	-.00594	.00011	-.00047	-.00424	-.00193	.01440

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-04-98 02:33:53 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	4.9643	4.9562	4.8413	5.0300	4.8113	5.1307
2	ICB	.00278	-.00517	.00882	.00369	.03292	.00162
3	LLS	.39728	.20072	.20908	.10486	.23059	Q.13473
4	ICSA	.01132	.01186	-.03289	-.00129	.04349	-.00619
5	ICSAB	.01974	.97514	-.03901	.00700	.05280	-.00760
6	S98T002821	635.20	59.163	1471.4	1.5768	59.895	202.33
7	S98T002821_D	635.53	66.566	1488.2	-.14636	54.141	198.46
8	S98T002961	358.08	79.599	728.03	-.52021	64.826	145.16
9	S98T002961_D	340.99	69.090	709.71	.07981	51.354	136.70
10	CCV	4.9947	5.0250	4.8783	5.0849	4.8376	5.1948
11	CCB	.00517	-.00245	.00707	.00110	.02868	.00989
12	S98T002962_L	335.09	27.503	718.61	-5.4968	115.63	181.51
13	S98T002962	335.20	105.32	723.70	.82425	55.050	179.50
14	S98T002962_D	317.39	85.263	690.62	1.2170	55.850	163.34
15	S98T002962_A	916.00	677.31	1296.2	602.85	658.57	783.15
16	S98T002962_X	349.94	73.458	772.51	-9.9648	229.24	196.86
17	S98T002962_AX	588.10	60194.	58981.	59533.	57777.	59802.
18	ICSA	.03619	.02945	-.02579	-.00339	.07722	-.00146
19	ICSAB	.04365	.98794	-.02929	.00226	.06604	-.00338
20	CCV	4.9870	5.0018	4.8416	5.0292	4.8076	5.1411
21	CCB	.00935	-.01584	-.00248	-.00123	.02945	.00262

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.9143	4.9275	.22451	5.0835	4.7707	9.7181
2	ICB	-.00140	.00000	.00415	.00024	-.00953	-.00387
3	LLS	.19586	.01993	.00775	.02013	.41870	.47732
4	ICSA	-.01827	.00195	-.00497	.00110	-.02470	-.03615
5	ICSAB	-.02134	.00191	.00063	.00112	.00029	-.04790
6	S98T002821	-3.6048	.59151	-1.8034	.25960	-10.143	-1.4783
7	S98T002821_D	-.83090	.59160	-3.1657	.52894	4.4359	-.41099
8	S98T002961	-.85266	.66669	2.5550	.54134	2.9876	17.701
9	S98T002961_D	-1.3762	.64186	-.88510	-.28178	-3.9525	10.588
10	CCV	4.9005	4.9273	.22455	5.0999	4.7836	9.6831
11	CCB	.00382	.00008	-.00435	.00020	-.00953	.02043
12	S98T002962_L	-14.843	.38848	-7.2442	-.71404	-9.5497	-24.828
13	S98T002962	.08432	.71877	-1.0095	.39734	2.5666	10.596
14	S98T002962_D	-.87886	.64186	-3.8731	.38509	-3.5439	10.518
15	S98T002962_A	573.35	576.92	20.776	594.61	552.56	1134.6
16	S98T002962_X	34.651	.75975	-26.303	1.2320	-6.6726	144.47
17	S98T002962_AX	57052.	58452.	2690.3	58387.	58428.	110960.
18	ICSA	-.01473	.00185	.00657	.00137	.01160	-.03659
19	ICSAB	-.01549	.00189	.00568	.00114	-.00154	-.04391
20	CCV	4.8349	4.8576	.22702	5.0336	4.7147	9.5593
21	CCB	.00112	.00004	-.00332	-.00001	.01125	.00778

HNF-1663 REV. 0

Analysis Report

Averages

Fri 12-04-98 02:33:53 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9633	.00746	4.6684	4.8410
2	ICB	.00034	-.00015	-.00014	-.00061
3	LLS	.10025	.00026	.01885	.01886
4	ICSA	.00080	.00714	-.00465	-.00404
5	ICSAB	.46197	.00699	.91687	-.00424
6	S98T002821	.27068	-.09357	.20875	-.25537
7	S98T002821_D	.27099	-.09402	.03550	.21252
8	S98T002961	.54126	-.05462	-.97232	-.04242
9	S98T002961_D	.25114	-.14579	-.66240	.03932
10	CCV	5.0001	.00787	4.7357	4.8709
11	CCB	.00189	.00022	-.00031	.00091
12	S98T002962_L	-2.0800	-1.1669	-4.4263	-1.2812
13	S98T002962	.45355	.04259	-1.5669	.41112
14	S98T002962_D	.15182	-.05156	-1.6334	.44300
15	S98T002962_A	585.39	.88367	563.67	577.76
16	S98T002962_X	10.257	1.3516	-7.0706	5.4798
17	S98T002962_AX	58475.	96.496	56100.	58113.
18	ICSA	.00221	.00753	-.00402	-.00337
19	ICSAB	.46336	.00746	.92339	-.00372
20	CCV	4.9410	.00808	4.7098	4.7989
21	CCB	.00069	.00007	-.00022	.00024

JK

12-04-98

File #: 981122G.7XT

HNF-1663 REV. 0

10/29/98 10:50

Page: 1

A-0004-1

LABCORE Data Entry Template for Worklist# 26892

Analyst: NJG/0 Instrument: ICPOL 2 11/03/98 Book# 77B30Method: LA-505-151/161 Rev/Mod C3Ensure dose rate at 30cm is 530 mrem/hr
prior to performing this analysisWorklist Comment: 11/01/98 ICP BY-105 (DIRECT)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SERDIL	S98T002963 0 D	@ICP-D01	LIQUID		
7	SAMPLE	S98T002963 0 D	@ICP-D01	LIQUID	98000509	BY-105 (2)
Analytes Requested:			AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01			
8	DUP	S98T002963 0 D	@ICP-D01	LIQUID		
9	SPK	(1 ppm) S98T002963 0 D	@ICP-D01	LIQUID		
10	CCV	Spiked GR (10 ppm) 11-11-98	@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SAMPLE	S98T002964 0 D	@ICP-D01	LIQUID	98000509	BY-105 (2)
Analytes Requested:			AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01			

Data Entry Comments:

Upgraded 12-29-98
John Wauell

Validated by:

Sam M. Pang
12/29/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1873

LABCORE Data Entry Template for Worklist# 26892

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 DUP	S98T002964	0 D	@ICP-D01	LIQUID		
14 ICSA			@ICP-QC	QC		
15 ICSAB			@ICP-QC	QC		
16 CCV			@ICP-QC	QC		
17 CCB			@ICP-QC	QC		

Final page for worklist # 26892

JK
 Analyst Signature   Date
 S98T002963-1 , 015-15-2-8 , DF 3005
 S98T002963 , 025-15 , 601
 S98T002963-D , 025-15 , 601
 S98T002963-Q , 025-15 , 601
 S98T002963-X , 025-15-1-9 , 6010
 S98T002963-QX , 025-15-1-9 , 6010
 S98T002964 , 025-15 , DF 601
 S98T002964-D , 025-15 , 601

Analyst Signature   Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

Summary

Tue 12-22-98 10:03:54 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981222A	ICP2	12/22/98	08:45	DKS	Q	CONC
2	ICB	981222A	ICP2	12/22/98	08:49	DKS	Q	CONC
3	LLS	981222A	ICP2	12/22/98	08:51	DKS	Q	CONC
4	ICSA	981222A	ICP2	12/22/98	08:54	DKS	Q	CONC
5	ICSAB	981222A	ICP2	12/22/98	08:57	DKS	Q	CONC
6	S98T002963_L	981222A	ICP2	12/22/98	09:01	DKS	S	CONC
7	S98T002963	981222A	ICP2	12/22/98	09:04	DKS	S	CONC
8	S98T002963_D	981222A	ICP2	12/22/98	09:07	DKS	S	CONC
9	S98T002963_A	981222A	ICP2	12/22/98	09:11	DKS	S	CONC
10	S98T002963_X	981222A	ICP2	12/22/98	09:18	DKS	S	CONC
11	S98T002963_AX	981222A	ICP2	12/22/98	09:21	DKS	S	CONC
12	CCV	981222A	ICP2	12/22/98	09:35	DKS	Q	CONC
13	CCB	981222A	ICP2	12/22/98	09:38	DKS	Q	CONC
14	S98T002964	981222A	ICP2	12/22/98	09:41	DKS	S	CONC
15	S98T002964_D	981222A	ICP2	12/22/98	09:44	DKS	S	CONC
16	ICSA	981222A	ICP2	12/22/98	09:51	DKS	Q	CONC
17	ICSAB	981222A	ICP2	12/22/98	09:54	DKS	Q	CONC
18	CCV	981222A	ICP2	12/22/98	09:57	DKS	Q	CONC
19	CCB	981222A	ICP2	12/22/98	10:00	DKS	Q	CONC

12-22-98

Worklist # 24892

BY-105

598T002963

598T002964

S98T002963

$$C_r = \frac{\left(\frac{1764.0}{601}\right) - \left(\frac{1269.5}{601}\right)}{1} \times 100 = 82.3\%$$

Post Spike:

$$R_L = \frac{\left(\frac{104240}{6010}\right) - \left(\frac{47349}{6010}\right)}{10} \times 100 = 94.7\%$$

$$N_a = \frac{\left(\frac{270260}{6010}\right) - \left(\frac{213280}{6010}\right)}{10} \times 100 = 94.8\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1875 TO 1876.

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9561	4.8936	5.0422	5.1204	4.8494	5.0779
2	ICB	-.00159	-.00463	-.01286	.00922	.00006	.00008
3	LLS	.02041	.10813	.20249	.10193	.09842	.01027
4	ICSA	.00159	243.81	.03201	.00093	.00020	.00021
5	ICSAB	.94123	241.57	.04020	-.00258	.45912	.47300
6	S98T002963_L	14.324	47805.	-1.8094	54.043	.19950	1.0581
7	S98T002963	16.187	47235.	-8.5246	34.647	.09794	.60647
8	S98T002963_D	14.369	44038.	-13.960	31.321	.01414	.50207
9	S98T002963_A	563.05	44459.	589.37	630.30	557.11	602.03
10	S98T002963_X	22.420	47349.	-60.475	30.497	.39696	1.8452
11	S98T002963_AX	57375.	104240.	59791.	59040.	59267.	58941.
12	CCV	4.9381	4.8749	5.0639	5.1384	4.9077	5.1297
13	CCB	.00288	.01454	-.00643	.00460	.00012	.00022
14	S98T002964	16.019	45845.	-22.869	26.887	.09988	.60695
15	S98T002964_D	15.882	46116.	-16.832	27.442	.20386	.65927
16	ICSA	-.00037	235.39	.06627	-.00129	.00008	.00052
17	ICSAB	.92686	235.05	.06359	.00018	.44816	.46577
18	CCV	4.8994	4.8133	5.0151	5.0683	4.8115	5.0764
19	CCB	-.00036	.00801	-.00570	.00737	.00000	.00030

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0162	5.1636	5.1014	4.9972	4.9793	4.9873
2	ICB	-.00711	-.00225	-.00108	-.01387	-.00138	-.00208
3	LLS	.18412	.23534	.00895	.19600	.04012	.01896
4	ICSA	-.01023	253.48	.00293	.00295	-.00103	-.00332
5	ICSAB	-.02164	252.26	.93554	-.00071	.45868	.46390
6	S98T002963_L	-42.024	18.457	-2.2941	-26.787	5.8787	1292.0
7	S98T002963	-9.6347	36.851	.79284	1.9775	4.4209	1269.5
8	S98T002963_D	.24987	10.084	-.13498	-4.5771	3.3841	1183.0
9	S98T002963_A	572.54	626.39	601.90	586.87	586.74	1764.0
10	S98T002963_X	-19.396	37.798	.46651	5.7980	-.68628	1285.4
11	S98T002963_AX	60194.	60145.	59276.	59119.	58945.	60555.
12	CCV	5.0374	5.1571	5.1128	5.0331	4.9794	4.9954
13	CCB	-.02214	.00161	.00067	-.00137	.00182	-.00005
14	S98T002964	-7.2531	11.811	.47239	1.5798	3.8693	1308.8
15	S98T002964_D	-2.8977	15.400	.72996	2.1760	4.4235	1323.8
16	ICSA	.00360	248.70	.00208	.00035	.00092	-.00121
17	ICSAB	.00800	250.68	.92428	.00608	.44926	.46025
18	CCV	5.0267	5.1632	5.0712	4.9427	4.9460	4.9602
19	CCB	-.01772	-.00111	-.00042	-.00726	.00034	-.00041

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7223	.00818	4.8210	4.7233	5.0175	4.9925
2	ICB	-.00085	-.00020	-.00087	-.27476	-.00178	.00021
3	LLS	.01645	-.00057	.09920	Q.11570	.10100	.02034
4	ICSA	.00160	.04207	93.298	.03008	-.00443	.00271
5	ICSAB	.45786	.04430	92.319	-.20748	-.00513	.98707
6	S98T002963_L	-1.1317	.61320	48.166	4819.2	-3.3483	.98254
7	S98T002963	1.5799	.16109	44.294	4789.1	.50380	.32748

Analysis Report

Averages

Tue 12-22-98 10:03:54 AM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
8	S98T002963_D	.62850	-.14131	40.697	4437.6	-.53803	.00039
9	S98T002963_A	549.70	1.0083	601.79	5060.2	592.11	570.45
10	S98T002963_X	4.3371	-.01267	38.240	3952.2	-.10835	3.9292
11	S98T002963_AX	55266.	97.643	56630.	63875.	59052.	57584.
12	CCV	4.7519	.00728	4.8101	4.8281	5.0636	4.9435
13	CCB	.00085	-.00106	.00112	-.20882	.00045	-.00065
14	S98T002964	.86597	.06301	38.998	4773.8	.45844	.58934
15	S98T002964_D	1.4890	.12243	40.400	4764.0	.87987	.71989
16	ICSA	.00234	.04122	90.915	.01714	-.00635	.00271
17	ICSAB	.44394	.03998	90.948	.05194	-.00348	.94153
18	CCV	4.6586	.00889	4.7467	4.5445	4.9687	4.8170
19	CCB	-.00001	.00001	-.00053	-.14702	-.00067	.00021

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1694	4.9320	5.0429	4.9446	4.8918	4.9358
2	ICB	-.01143	.00009	.00186	-.01150	-.00784	-.00596
3	LLS	.20937	.01994	.10105	.20441	.19374	.03613
4	ICSA	253.20	-.00568	-.01278	194.05	-.00209	-.00454
5	ICSAB	250.06	.43901	-.01331	192.30	.00056	.91521
6	S98T002963_L	-13.286	.47150	33.519	217400.	-12.045	-6.8221
7	S98T002963	-.86160	.03931	36.757	210590.	.41145	-1.6256
8	S98T002963_D	-1.1164	.01660	33.036	196340.	-2.3881	-2.6031
9	S98T002963_A	598.78	566.37	632.54	195460.	574.22	578.53
10	S98T002963_X	12.407	.39525	37.253	213280.	-18.981	-16.870
11	S98T002963_AX	60712.	59043.	58875.	270260.	57202.	59256.
12	CCV	5.1400	4.9154	5.0528	4.9119	4.9068	4.9615
13	CCB	-.00072	-.00015	.00168	.00963	.00002	-.00400
14	S98T002964	2.0583	.02099	36.759	209170.	-.12393	-.11732
15	S98T002964_D	6.2452	.19777	34.156	210340.	-.62572	1.6433
16	ICSA	244.25	-.00534	-.01389	184.55	.00195	-.00400
17	ICSAB	243.72	.43320	-.01363	184.19	-.00093	.90059
18	CCV	5.0585	4.8723	4.9995	4.7779	4.7904	4.9425
19	CCB	-.00945	.00014	.00031	-.00707	-.00307	-.00074

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0317	5.0157	4.8733	5.0956	4.8196	5.2006
2	ICB	-.00165	-.00869	.00803	-.00997	.00589	.00340
3	LLS	.39726	.20049	.20422	.09407	.18435	Q.24228
4	ICSA	.02188	.02758	-.01813	.01611	.03156	-.00757
5	ICSAB	.03442	.98542	-.02177	.00496	.04245	.00119
6	S98T002963_L	429.47	104.26	755.42	-9.9073	82.705	277.95
7	S98T002963	448.46	97.031	747.90	2.7003	63.928	272.81
8	S98T002963_D	421.80	84.732	693.48	-2.4129	54.536	255.04
9	S98T002963_A	1014.6	672.26	1262.8	605.36	650.17	880.10
10	S98T002963_X	446.46	93.928	766.34	-18.322	156.85	328.26
11	S98T002963_AX	58736.	59726.	58408.	59800.	57458.	60890.
12	CCV	5.0110	5.0116	4.8381	5.1257	4.8380	5.1762
13	CCB	.01113	.00198	.00940	-.00310	.01574	.00330
14	S98T002964	368.21	92.039	721.76	1.1669	55.557	206.84
15	S98T002964_D	406.65	84.354	727.30	-.30283	60.200	214.60

Analysis Report

Averages

Tue 12-22-98 10:03:54 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
16	ICSA	.03466	.03242	-.02776	.00085	.06983	.00114
17	IC SAB	.04623	.08891	-.01164	.00054	.07668	.00494
18	CCV	4.9287	5.0014	4.8162	5.0901	4.8056	5.1388
19	CCB	.00985	-.00450	.01026	.00431	.03169	.00628

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8704	4.8967	.20482	5.0887	4.7392	9.6493
2	ICB	-.00352	.00000	-.01023	-.00114	-.01582	-.01027
3	LLS	.18753	.01995	.00543	.01884	.39600	.44177
4	ICSA	-.02053	.00188	.00763	-.00005	-.03090	.10232
5	IC SAB	-.01842	.00188	.00882	.00083	-.04907	.10673
6	S98T002963_L	-4.4719	.65484	-13.415	-3.3733	-53.886	-6.7671
7	S98T002963	-3.4617	.62756	-2.7121	.25106	-11.127	-1.7576
8	S98T002963_D	-4.3849	.55039	-3.8837	-.28297	-13.126	-7.3129
9	S98T002963_A	571.83	575.32	18.531	595.28	535.99	1126.8
10	S98T002963_X	-10.338	.52336	-2.5385	-5.3311	-168.69	-8.8183
11	S98T002963_AX	5741.4	58418.	2593.3	58593.	58184.	111260.
12	CCV	4.9325	4.9328	.22273	5.1207	4.7801	9.6960
13	CCB	-.01144	-.00012	.00285	.00001	-.01447	-.03836
14	S98T002964	-2.1476	.57595	-1.7335	-.14212	-18.826	-.72652
15	S98T002964_D	-1.5186	.60136	-.36219	-.13982	-16.206	6.7297
16	ICSA	-.01337	.00174	.01766	.00044	-.07009	.07940
17	IC SAB	-.01009	.00185	.01877	.00045	-.03432	.08878
18	CCV	4.8294	4.8443	.22510	5.0548	4.7369	9.5033
19	CCB	-.00140	-.00004	-.00182	-.00067	-.02463	.00010

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9899	.00757	4.7430	4.8554
2	ICB	-.00063	-.00038	-.00034	.00047
3	LLS	.09734	-.00018	.01941	.01761
4	ICSA	-.00008	.00716	-.00423	-.00445
5	IC SAB	.46181	.00732	.92312	-.00428
6	S98T002963_L	.50394	-.46827	-1.8339	.08278
7	S98T002963	-.25279	-.23482	-1.5988	.54418
8	S98T002963_D	-.43630	-.37381	-1.6280	.27907
9	S98T002963_A	585.47	.88743	561.12	575.30
10	S98T002963_X	2.0016	-.46741	2.4168	2.0289
11	S98T002963_AX	5841.4	92.177	55506.	58228.
12	CCV	4.9970	.00732	4.7062	4.8768
13	CCB	-.00180	-.00053	.00025	-.00168
14	S98T002964	.15185	-.23526	2.9671	.28810
15	S98T002964_D	.34126	-.14420	3.1625	.18139
16	ICSA	.00153	.00725	-.00461	-.00437
17	IC SAB	.45724	.00772	.91290	-.00369
18	CCV	4.9466	.00772	4.6929	4.8098
19	CCB	.00051	-.00007	.00044	.00014

Handwritten signature
12-22-98

LABCORE Data Entry Template for Worklist# 26893

Analyst: BK Seto Instrument: ICPOM 2 11-02-98 Book# 77B30Method: LA-505-151/161 Rev/Mod C-3Worklist Comment: ICP BY-105 (DIRECT)Ensure dose rate at 30 cm is ≤ 50 μ rem/hr
prior to performing this analysis.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SAMPLE	S98T003006 0 D	@ICP-D01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01					
7	DUP	S98T003006 0 D	@ICP-D01	LIQUID		
8	SAMPLE	S98T003007 0 D	@ICP-D01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PB-D-01, S-D-01, SB-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01					
9	DUP	S98T003007 0 D	@ICP-D01	LIQUID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T003008 0 D	@ICP-D01	LIQUID		

Data Entry Comments:

uploaded 12-14-98
John Wainell

Validated by:

Saul M. Pang
12/15/97

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26893

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 SAMPLE	S98T003008	0 D	@ICP-D01	LIQUID	98000510	BY-105 (2)
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
14 DUP	S98T003008	0 D	@ICP-D01	LIQUID		
15 SPK (1 ppm) Spike X	S98T003008	0 D	@ICP-D01	LIQUID		
16 SAMPLE	S98T003009	0 D	@ICP-D01	LIQUID	98000510	BY-105 (2)
Analytes Requested: AG-D-01 , AL-D-01 , AS-D-01 , B-D-01 , BA-D-01 , BE-D-01 , BI-D-01 , CA-D-01 , CD-D-01 , CE-D-01 , CO-D-01 , CR-D-01 , CU-D-01 , FE-D-01 , K-D-01 , LA-D-01 , LI-D-01 , MG-D-01 , MN-D-01 , MO-D-01 , NA-D-01 , ND-D-01 , NI-D-01 , P-D-01 , PB-D-01 , S-D-01 , SB-D-01 , SE-D-01 , SI-D-01 , SM-D-01 , SR-D-01 , TI-D-01 , TL-D-01 , U-D-01 , V-D-01 , ZN-D-01 , ZR-D-01						
17 DUP	S98T003009	0 D	@ICP-D01	LIQUID		
18 ICSA			@ICP-QC	QC		
19 ICSAB			@ICP-QC	QC		
20 CCV			@ICP-QC	QC		
21 CCB			@ICP-QC	QC		

Final page for worklist # 26893

JK 12-14-98
 Analyst Signature Date
 S98T003006 , 025-15 , DF 601
 S98T003006 , 025-15 , 601
 S98T003007 , 025-15 , DF 601
 S98T003007 , 025-15 , 601
 Data Entry Comments: S98T003009 , 025-15 , DF 601
 S98T003009 , 025-15 , 601

Analyst Signature Date
 S98T003008-L , 025-15-28 , JF 3005
 S98T003008 , 025-15 , 601
 S98T003008 , 025-15 , 601
 S98T003008 , 025-15 , 601
 S98T003008-X , 025-15-19 , 6010
 S98T003008-QX , 025-15-19 , 6016

Analysis Report

Summary

Mon 12-14-98 02:25:31 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981214C	ICP2	12/14/98	12:41	DKS	Q	CONC
2	ICB	981214C	ICP2	12/14/98	12:44	DKS	Q	CONC
3	LLS	981214C	ICP2	12/14/98	12:47	DKS	Q	CONC
4	ICSA	981214C	ICP2	12/14/98	12:50	DKS	Q	CONC
5	ICSAB	981214C	ICP2	12/14/98	12:53	DKS	Q	CONC
6	S98T003006	981214C	ICP2	12/14/98	12:57	DKS	S	CONC
7	S98T003006_D	981214C	ICP2	12/14/98	13:00	DKS	S	CONC
8	S98T003007	981214C	ICP2	12/14/98	13:04	DKS	S	CONC
9	S98T003007_D	981214C	ICP2	12/14/98	13:07	DKS	S	CONC
10	CCV	981214C	ICP2	12/14/98	13:12	DKS	Q	CONC
11	CCB	981214C	ICP2	12/14/98	13:15	DKS	Q	CONC
12	S98T003008_L	981214C	ICP2	12/14/98	13:18	DKS	S	CONC
13	S98T003008	981214C	ICP2	12/14/98	13:22	DKS	S	CONC
14	S98T003008_D	981214C	ICP2	12/14/98	13:25	DKS	S	CONC
15	S98T003008_A	981214C	ICP2	12/14/98	13:28	DKS	S	CONC
16	S98T003008_X	981214C	ICP2	12/14/98	13:37	DKS	S	CONC
17	S98T003008_AX	981214C	ICP2	12/14/98	13:40	DKS	S	CONC
18	S98T003009	981214C	ICP2	12/14/98	14:00	DKS	S	CONC
19	S98T003009_D	981214C	ICP2	12/14/98	14:04	DKS	S	CONC
20	ICSA	981214C	ICP2	12/14/98	14:12	DKS	Q	CONC
21	ICSAB	981214C	ICP2	12/14/98	14:15	DKS	Q	CONC
22	CCV	981214C	ICP2	12/14/98	14:18	DKS	Q	CONC
23	CCB	981214C	ICP2	12/14/98	14:22	DKS	Q	CONC

JK

BY-105

12-14-98

Workfile # 16893

598T003006

598T003007

598T003008

598T003009

598T003008

Post Spike:

$$Ql = \frac{\left(\frac{103070}{6010}\right) - \left(\frac{46050}{6010}\right)}{10} \times 100 = 94.9\%$$

$$Na = \frac{\left(\frac{272740}{6010}\right) - \left(\frac{215970}{6010}\right)}{10} \times 100 = 94.46\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1001 TO 1005

Analysis Report

Averages

Mon 12-14-98 02:25:31 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9655	4.7547	4.9774	5.0581	4.7325	4.9818
2	ICB	-.00149	-.01023	.00198	-.00504	-.00002	-.00008
3	LLS	.02050	.09223	.19751	.12021	.09728	.01004
4	ICSA	.00064	237.67	.02274	-.01060	.00029	.00013
5	ICSAB	.94906	239.68	.01567	-.01068	.45409	.47135
6	S98T003006	13.919	37996.	-13.110	25.956	-.01334	.41500
7	S98T003006_D	13.954	37427.	-10.321	23.747	.02650	.38883
8	S98T003007	14.723	37958.	-14.401	19.331	.12386	.38846
9	S98T003007_D	13.716	37436.	-6.3104	20.990	.04304	.33698
10	CCV	4.9707	4.7626	5.0174	5.0696	4.7345	5.0262
11	CCB	-.000196	-.00848	-.00056	.00046	.00005	.00012
12	S98T003008_L	11.818	44924.	-17.563	33.172	.17246	.64754
13	S98T003008	13.796	44595.	-22.985	24.305	.11130	.36309
14	S98T003008_D	13.969	43791.	-11.020	24.580	.16733	.46628
15	S98T003008_A	572.83	44903.	584.40	610.99	542.71	592.31
16	S98T003008_X	7.8137	46050.	-25.685	5.5316	-.37569	.26667
17	S98T003008_AX	58786.	103070.	60681.	59504.	59395.	59211.
18	S98T003009	13.579	43141.	-5.7583	25.404	.04189	.46691
19	S98T003009_D	13.042	42768.	-7.3878	27.616	.01050	.38899
20	ICSA	.00024	240.01	.03772	-.01340	.00023	.00004
21	ICSAB	.94973	238.34	.03224	-.01161	.45312	.46963
22	CCV	5.0116	4.7736	5.0421	5.1040	4.7512	5.0277
23	CCB	-.00147	-.01062	.00476	.00001	.00005	-.00012

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9798	5.1705	5.0693	4.9146	4.9621	4.9807
2	ICB	-.00713	.00795	-.00126	-.00596	-.00058	-.00040
3	LLS	.21301	.23267	.00890	.19274	.03585	.01889
4	ICSA	-.00752	252.34	.00464	-.00299	.00045	-.00288
5	ICSAB	-.03063	255.70	.93866	-.00183	.46153	.46420
6	S98T003006	-1.7194	6.1431	-1.11069	-5.4866	3.4579	1289.7
7	S98T003006_D	-4.1189	5.4899	-.62614	-4.1420	4.1773	1272.0
8	S98T003007	1.4928	8.8945	.06150	-2.8019	4.2470	1264.8
9	S98T003007_D	-1.2019	7.0501	-1.3007	-6.8143	1.8033	1244.1
10	CCV	5.0219	5.1897	5.0950	4.9111	4.9839	5.0053
11	CCB	-.01756	.00919	-.00107	-.00402	.00026	-.00025
12	S98T003008_L	-44.682	16.739	-2.2993	-12.073	1.1655	1336.9
13	S98T003008	-8.6018	37.000	.00646	-2.9834	3.6810	1311.3
14	S98T003008_D	-15.114	16.253	-.54389	-1.0912	3.5342	1286.6
15	S98T003008_A	579.89	631.26	602.75	570.59	588.50	1888.2
16	S98T003008_X	-113.56	-21.949	-3.4235	-70.181	5.0639	1351.1
17	S98T003008_AX	60621.	61104.	60085.	59729.	59651.	61645.
18	S98T003009	-14.086	12.262	-.50208	-7.5747	3.3827	1389.9
19	S98T003009_D	-10.665	7.1607	-.24812	-8.3062	3.2468	1390.0
20	ICSA	.00504	255.57	.00347	-.00376	-.00218	-.00477
21	ICSAB	-.04137	256.22	.94098	-.00336	.46021	.46829
22	CCV	4.9976	5.2735	5.1337	4.9200	5.0297	5.0476
23	CCB	.01936	.00833	-.00202	-.00466	-.00164	-.00105

Analysis Report

Averages

Mon 12-14-98 02:25:31 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.6197	.00933	4.7729	4.6601	4.9330	4.7979
2	ICB	-.00060	.00065	-.00084	-.13916	-.00036	.00030
3	LLS	.01623	.00029	.09677	Q.26452	.10025	.02023
4	ICSA	.00258	.02904	91.537	-.11065	-.00435	.00201
5	ICSAB	.45797	.03105	92.451	-.00157	-.00419	.96678
6	S98T003006	-.16926	.70035	33.121	4251.3	-.61754	1.6350
7	S98T003006_D	.22408	.86248	32.674	4146.7	-.66900	1.8770
8	S98T003007	.36551	.77421	32.838	4343.3	.11980	1.5740
9	S98T003007_D	-.58562	.64811	31.900	4228.5	-.58908	1.4532
10	CCV	4.6277	.00971	4.7873	4.6321	4.9377	4.8080
11	CCB	-.00115	.00184	-.00087	-.11231	-.00040	.00140
12	S98T003008_L	-1.9677	5.9101	40.162	3945.9	-.67653	5.1399
13	S98T003008	1.5746	1.2551	40.556	4406.9	-.28232	2.4218
14	S98T003008_D	1.7257	1.0783	39.547	4430.0	.35071	2.1798
15	S98T003008_A	540.26	1.8905	600.11	5012.0	581.28	554.01
16	S98T003008_X	-1.9880	1.9647	34.889	3980.0	-4.4007	3.0276
17	S98T003008_AX	55340.	107.53	57496.	63193.	59433.	57882.
18	S98T003009	.43469	.58775	34.012	4384.8	-.51354	1.2715
19	S98T003009_D	.01502	1.1445	33.708	4414.1	-.47575	1.7552
20	ICSA	.00210	.02626	92.670	-.13189	-.00466	.00232
21	ICSAB	.45479	.02632	92.394	.00289	-.00463	.96849
22	CCV	4.6243	.01058	4.8033	4.9451	4.9319	4.8422
23	CCB	-.00101	.00153	-.00117	-.11431	-.00007	.00080

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.0413	4.8933	5.0268	4.7617	4.8044	4.9724
2	ICB	-.00510	.00016	.00301	-.01795	-.00034	-.00494
3	LLS	.19930	.02011	.10070	.17700	.19656	.03893
4	ICSA	245.07	-.00491	-.01329	188.13	-.00231	-.00697
5	ICSAB	246.64	.44114	-.01225	188.64	-.00345	.90856
6	S98T003006	-8.6185	.16553	32.851	208580.	-2.1390	-2.6298
7	S98T003006_D	-6.2548	.19387	31.902	203790.	-.46315	-2.1589
8	S98T003007	-2.0248	.23248	32.469	206160.	1.3154	1.8716
9	S98T003007_D	-8.3551	.10889	32.757	204780.	-1.8475	-.27115
10	CCV	5.0373	4.9058	5.0594	4.7424	4.8012	4.9665
11	CCB	-.00525	.00055	.00271	-.01450	-.00103	-.00482
12	S98T003008_L	-9.3056	1.8175	32.859	207230.	-6.9887	-1.0613
13	S98T003008	-.42447	.48900	33.997	203490.	-1.2642	-1.2173
14	S98T003008_D	.42810	.40550	33.138	200680.	-.37750	-1.8282
15	S98T003008_A	584.17	567.46	636.14	203900.	562.85	588.52
16	S98T003008_X	-72.478	.11378	26.594	215970.	-1.0219	-43.775
17	S98T003008_AX	60873.	59886.	59836.	272740.	57949.	60318.
18	S98T003009	-10.430	.28748	34.651	207060.	-.77337	-2.0959
19	S98T003009_D	-10.184	.44634	35.802	203600.	-1.4857	-.20691
20	ICSA	247.49	-.00507	-.01550	189.39	-.00516	.00169
21	ICSAB	245.54	.44046	-.01682	187.69	-.00177	.92016
22	CCV	5.0458	4.9423	5.0917	4.7828	4.8045	5.0095
23	CCB	-.01060	.00033	-.00044	-.01415	-.00119	-.00549

Analysis Report

Averages

Mon 12-14-98 02:25:31 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	4.9494	5.0474	4.8756	5.0456	4.8519	5.1624
2	ICB	-.00095	-.00379	.00225	-.00898	.00998	.00365
3	LLS	.39835	.19504	.20083	.09053	.20235	Q.24485
4	ICSA	.02571	.02543	-.02188	-.00206	.06565	-.00124
5	ICSAB	.03712	.99789	-.02621	-.01028	.07501	.00008
6	S98T003006	426.26	68.405	826.79	-1.3359	49.720	161.60
7	S98T003006_D	391.19	77.334	819.37	-2.7179	60.399	155.09
8	S98T003007	383.41	82.409	816.76	-4.1765	59.983	130.94
9	S98T003007_D	363.85	67.523	805.62	-4.3524	57.738	126.47
10	CCV	5.0096	5.0688	4.9149	5.0813	4.8565	5.1912
11	CCB	.00391	-.00617	.00884	-.00664	.03433	.01253
12	S98T003008_L	446.33	31.798	803.54	-34.666	164.17	201.18
13	S98T003008	478.62	80.707	789.63	-7.1905	62.013	146.71
14	S98T003008_D	459.00	81.413	782.49	-6.3746	54.823	147.09
15	S98T003008_A	1085.8	681.51	1381.2	605.20	658.40	774.42
16	S98T003008_X	479.17	80.933	794.50	-57.473	195.43	221.66
17	S98T003008_AX	60030.	61092.	59605.	60005.	58833.	60046.
18	S98T003009	449.36	72.630	833.33	-9.0898	65.139	171.18
19	S98T003009_D	453.94	79.831	855.66	-4.4459	68.317	174.82
20	ICSA	.03513	.01423	-.02186	.00055	.07831	-.00153
21	ICSAB	.04771	1.0047	-.03677	.00231	.07715	.00415
22	CCV	5.0673	5.1146	4.9493	5.1044	4.8686	5.2236
23	CCB	.00977	-.00787	.00550	-.00459	.02198	.00788

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.7710	4.8067	.23509	5.0108	4.7370	9.4578
2	ICB	.00620	.00003	-.00767	-.00090	.00957	.02143
3	LLS	.19491	.01985	.00015	.01883	.39981	.47229
4	ICSA	-.01961	.00172	.00202	.00088	-.01143	-.05427
5	ICSAB	-.01701	.00173	.00500	.00068	-.01044	-.06447
6	S98T003006	4.1642	.71947	-7.8993	-4.2744	-3.8297	18.924
7	S98T003006_D	4.7648	.74382	-7.2373	-.03335	-5.8898	25.146
8	S98T003007	4.7439	.76845	-5.8369	.36505	11.352	26.622
9	S98T003007_D	2.7707	.71934	-10.318	-.43928	-5.0428	20.134
10	CCV	4.7699	4.8157	.23556	5.0381	4.7565	9.4649
11	CCB	.01711	.00024	-.00430	.00019	-.01335	.05720
12	S98T003008_L	50.930	1.2260	-29.870	-2.1079	.72774	194.98
13	S98T003008	9.4408	.69203	-7.6698	-.42750	-6.9521	35.864
14	S98T003008_D	7.8111	.66840	-5.2867	-.02392	1.6839	29.911
15	S98T003008_A	563.75	563.91	21.328	586.96	555.73	1116.7
16	S98T003008_X	9.4080	.49479	-52.493	-.20562	-42.984	39.400
17	S98T003008_AX	57321.	58769.	2872.7	58983.	59216.	111380.
18	S98T003009	1.9925	.69451	-8.2422	.09394	3.9950	17.562
19	S98T003009_D	7.1297	.74275	-12.113	-.05431	-6.4986	36.714
20	ICSA	-.01884	.00165	.00795	.00090	-.05184	-.06118
21	ICSAB	-.01548	.00167	.00363	.00154	-.01605	-.05233
22	CCV	4.7598	4.8183	.23460	5.0289	4.7881	9.4499
23	CCB	.01267	.00011	-.01219	-.00092	-.01131	.04465

Analysis Report

Averages

Mon 12-14-98 02:25:31 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9586	.00851	4.7752	4.8073
2	ICB	.00150	.00022	-.00057	.00204
3	LLS	.09953	.00026	.01963	.01948
4	ICSA	.00079	.00746	-.00335	-.00324
5	ICSAB	.46542	.00778	.93446	-.00338
6	S98T003006	1.4657	.08615	-1.8618	1.8063
7	S98T003006_D	1.6580	.22481	-1.6662	1.7494
8	S98T003007	1.6501	.22418	-1.9809	1.6419
9	S98T003007_D	.95984	.13274	-2.0124	1.9861
10	CCV	4.9777	.00867	4.7947	4.8198
11	CCB	.00380	.00083	.00030	.00294
12	S98T003008_L	14.460	3.2142	-.12304	10.984
13	S98T003008	2.5522	.54996	-.52013	2.3061
14	S98T003008_D	2.1502	.41119	-.73508	1.9558
15	S98T003008_A	585.95	1.2682	574.96	573.44
16	S98T003008_X	5.1211	-.01674	-2.5562	9.2436
17	S98T003008_AX	59056.	99.146	56732.	58712.
18	S98T003009	1.2621	.03963	-1.9545	1.6523
19	S98T003009_D	2.2557	.40830	-1.9488	2.6382
20	ICSA	.00102	.00762	-.00380	-.00348
21	ICSAB	.46445	.00778	.93587	-.00268
22	CCV	5.0032	.00891	4.8528	4.8274
23	CCB	.00300	.00060	-.00021	.00325

12-14-98

LABCORE Data Entry Template for Worklist# 26896

Analyst: JK Jcto Instrument: ICP012 10-30-98Book# 75B48BMethod: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30 cm is ≤ 50 μ rad/hr
prior to performing this analysis.Worklist Comment: 10-30-98
ICP BY-105 (FUSION)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@ICP-QC	QC		
2 ICB			@ICP-QC	QC		
3 LLS			@ICP-QC	QC		
4 ICSA			@ICP-QC	QC		
5 ICSAB			@ICP-QC	QC		
6 PREPBLKTJA			@ICP-F01	SOLID		
7 SERDIL	S98T002598	0 F	@ICP-F01	SOLID		
8 SAMPLE	S98T002598	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
9 DUP	S98T002598	0 F	@ICP-F01	SOLID		
10 SPK, (1 ppm)	S98T002598	0 F	@ICP-F01	SOLID		
11 <u>Sample X</u> <u>8.166 g/L (10 ppm)</u> <u>10-30-98</u>			@ICP-QC	QC		
12 CCB			@ICP-QC	QC		
13 SAMPLE	S98T002599	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 ,						

Data Entry Comments:

uploaded 11-2-98
John Wauell

Validated by:

Sam R. Pang
11/2/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26896

S Type	Sample#	R A	Test	Matrix	Group#	Project
			SR-F-01 , ZR-F-01	TI-F-01 , TL-F-01	U-F-01 , V-F-01	ZN-F-01 ,
14 DUP	S98T002599	0 F	@ICP-F01	SOLID		
15 SAMPLE	S98T002600	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
	Analytes Requested:					
			AG-F-01 , BE-F-01 , CR-F-01 , MN-F-01 , PB-F-01 , SR-F-01 , ZR-F-01	AL-F-01 , BI-F-01 , CU-F-01 , MO-F-01 , S-F-01 , TI-F-01	AS-F-01 , CA-F-01 , FE-F-01 , NA-F-01 , SB-F-01 , U-F-01	B-F-01 , CD-F-01 , LA-F-01 , ND-F-01 , SE-F-01 , V-F-01
						BA-F-01 , CE-F-01 , CO-F-01 , MG-F-01 , NI-F-01 , P-F-01 , SM-F-01 , ZN-F-01
16 DUP	S98T002600	0 F	@ICP-F01	SOLID		
17 ICSA			@ICP-QC	QC		
18 ICSAB			@ICP-QC	QC		
19 CCV			@ICP-QC	QC		
20 CCB			@ICP-QC	QC		

Final page for worklist # 26896

JK *10/30/98*

Analyst Signature	Date
<i>PrpblkTJA</i>	<i>1.25-10</i>
<i>5987001598-L</i>	<i>1.25-10-2-8</i>
<i>5987001598</i>	<i>1.25-10</i>
<i>5987001598-D</i>	<i>1.25-10</i>
<i>5987001598-G</i>	<i>1.25-10</i>
<i>5987001598-X</i>	<i>1.25-10-1-9</i>
<i>5987001598-QX</i>	<i>1.25-10-1-9</i>
<i>5987001598</i>	<i>1.25-10</i>
<i>5987001598-D</i>	<i>1.25-10</i>
<i>5987001600</i>	<i>1.25-10</i>
<i>5987001600-D</i>	<i>1.25-10</i>

Analyst Signature	Date

Data Entry Comments:

$$\begin{aligned}
 \text{al} &= 1.1756 \frac{\mu\text{g}}{\text{ml}} \times \frac{1}{2.0372 \frac{\text{g}}{\text{L}} \times \frac{1\text{L}}{100\text{ml}}} = \\
 &= 5.77 \pm 2 \frac{\mu\text{g}}{\text{g}}
 \end{aligned}$$

Sp. ki:

$$\text{al} = \frac{\left(\frac{40.707}{41}\right) - \left(\frac{1.1756}{41}\right)}{1} \times 100 = 96.4\%$$

Analysis Report

Summary

Fri 10-30-98 06:36:45 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981030D	ICP2	10/30/98	17:08	DKS	Q	CONC
2	ICB	981030D	ICP2	10/30/98	17:11	DKS	Q	CONC
3	LLS	981030D	ICP2	10/30/98	17:14	DKS	Q	CONC
4	ICSA	981030D	ICP2	10/30/98	17:17	DKS	Q	CONC
5	ICSAB	981030D	ICP2	10/30/98	17:20	DKS	Q	CONC
6	PREPBLKTJA	981030D	ICP2	10/30/98	17:24	DKS	Q	CONC
7	S98T002598_L	981030D	ICP2	10/30/98	17:28	DKS	S	CONC
8	S98T002598	981030D	ICP2	10/30/98	17:31	DKS	S	CONC
9	S98T002598_D	981030D	ICP2	10/30/98	17:34	DKS	S	CONC
10	S98T002598_A	981030D	ICP2	10/30/98	17:37	DKS	S	CONC
11	S98T002598_X	981030D	ICP2	10/30/98	17:43	DKS	S	CONC
12	S98T002598_AX	981030D	ICP2	10/30/98	17:46	DKS	S	CONC
13	CCV	981030D	ICP2	10/30/98	18:02	DKS	Q	CONC
14	CCB	981030D	ICP2	10/30/98	18:05	DKS	Q	CONC
15	S98T002599	981030D	ICP2	10/30/98	18:09	DKS	S	CONC
16	S98T002599_D	981030D	ICP2	10/30/98	18:12	DKS	S	CONC
17	S98T002600	981030D	ICP2	10/30/98	18:15	DKS	S	CONC
18	S98T002600_D	981030D	ICP2	10/30/98	18:18	DKS	S	CONC
19	ICSA	981030D	ICP2	10/30/98	18:23	DKS	Q	CONC
20	ICSAB	981030D	ICP2	10/30/98	18:26	DKS	Q	CONC
21	CCV	981030D	ICP2	10/30/98	18:30	DKS	Q	CONC
22	CCB	981030D	ICP2	10/30/98	18:33	DKS	Q	CONC

JK LLD

10-30-98

Worklist # 26896

B 4-105

598T002598

598T002599

598T002600

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1889 TO 1894.

Analysis Report

Averages

Fri 10-30-98 06:36:45 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0012	4.9803	5.1709	4.9915	5.0109	5.1178
2	ICB	-.00013	-.00040	-.00339	.00430	.00006	.00004
3	LLS	.02042	.09907	.19521	.10004	.09816	.01037
4	ICSA	.00310	245.10	.05472	-.00602	.00031	.00009
5	ICSAB	.94029	241.88	.01751	-.00769	.45162	.47007
6	PREPELKTJA	-.00033	-.00447	.00788	-.00049	.00006	.00005
7	S98T002598_L	-.06921	-.15621	1.5496	.29299	-.00231	.00979
8	S98T002598	.00123	1.1756	-.45501	.01911	.00248	.00396
9	S98T002598_D	.05862	1.8009	-.27226	.01984	.00759	-.00002
10	S98T002598_A	38.054	40.707	41.782	39.542	38.723	40.344
11	S98T002598_X	-.17458	-2.3067	-.43698	.58671	-.01496	-.01883
12	S98T002598_AX	3839.0	3982.9	4113.9	4017.9	4086.4	4093.2
13	CCV	5.0161	4.9949	5.2107	5.0331	5.0527	5.1523
14	CCB	-.00001	-.00339	.00464	.00286	-.00004	.00014
15	S98T002599	.04533	6.7080	.05626	.11736	.00563	-.00171
16	S98T002599_D	.01382	6.9979	-.48184	-.09849	.00308	-.00359
17	S98T002600	.00911	32.919	-.07645	.03905	.00988	.00013
18	S98T002600_D	.04491	48.808	-.10968	.03931	.01267	.00203
19	ICSA	.00158	248.91	.02865	-.01278	.00013	.00009
20	ICSAB	.94688	243.57	.04788	-.00684	.45659	.47657
21	CCV	5.0726	5.0087	5.2018	5.0278	5.0454	5.1641
22	CCB	-.00066	-.00346	.00081	.00094	-.00005	-.00004

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9743	4.9505	5.0828	5.1147	4.9866	4.9340
2	ICB	-.00336	.00234	.00056	.00063	.00048	.00057
3	LLS	.19634	.21927	.00943	.19951	.04009	.01916
4	ICSA	-.01143	253.81	.00525	.00755	-.00113	-.00282
5	ICSAB	-.02978	252.95	.92850	.01432	.45869	.45835
6	PREPELKTJA	.00038	.04311	.00017	.00435	-.00068	-.00075
7	S98T002598_L	-.86884	4.0638	-.25509	-.57622	-.12949	-.28610
8	S98T002598	-.36205	1.1674	-.02888	-.06019	.00370	.05096
9	S98T002598_D	.02400	-.03120	.01902	.29454	.04559	.11973
10	S98T002598_A	38.309	40.600	41.099	40.335	40.134	39.936
11	S98T002598_X	-3.2284	-2.1368	-.21772	-1.2663	.45448	-.15740
12	S98T002598_AX	4129.8	4118.8	4083.9	4112.0	4041.0	4084.5
13	CCV	5.0635	4.9789	5.1065	5.1393	4.9985	4.9600
14	CCB	.01244	-.00189	.00014	-.00280	.00124	-.00147
15	S98T002599	-.38051	-.09200	.00726	.08082	-.02139	.22246
16	S98T002599_D	-.84105	-.04997	-.00188	-.21230	-.02634	.15781
17	S98T002600	-.92806	.14793	-.02886	.03925	.04455	.96322
18	S98T002600_D	-.09110	.33114	-.01759	.19261	-.02148	.91781
19	ICSA	-.02210	255.70	.00427	-.00360	.00046	-.00422
20	ICSAB	-.01136	255.95	.93976	.01043	.46128	.46572
21	CCV	5.0199	5.0690	5.1474	5.1552	5.0469	5.0212
22	CCB	-.00384	-.00086	-.00060	-.00212	-.00038	-.00245

Analysis Report

Averages

Fri 10-30-98 06:36:45 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.8203	.00480	4.9150	5.0835	5.0715	5.1039
2	ICB	-.00010	-.00066	.00049	.04418	.00062	-.00032
3	LLS	.01728	-.00047	.09796	.51861	.01095	.01985
4	ICSA	.00286	.01253	93.355	.02183	-.00372	.00198
5	ICSAB	.45279	.01193	92.419	.20400	-.00239	.96030
6	PREPBLKTJA	.00038	.00005	.00319	Q270.70	.00043	.00011
7	S98T002598_L	-.11642	-.15360	.75128	11115.	-.03306	-.20321
8	S98T002598	.01700	-.02269	.72427	10856.	-.01969	-.01788
9	S98T002598_D	.03723	.01433	.69030	9724.5	.04197	.02714
10	S98T002598_A	37.912	.05387	40.097	10973.	39.785	39.006
11	S98T002598_X	.01148	-.07133	.85209	10991.	-.16452	.13620
12	S98T002598_AX	3816.0	4.6604	3919.4	15053.	4098.4	4050.8
13	CCV	4.8487	.00474	4.9379	4.9312	5.1035	5.1434
14	CCB	-.00035	-.00104	.00069	-.06835	-.00058	-.00099
15	S98T002599	.03037	.00059	.12756	10576.	.04323	.00017
16	S98T002599_D	-.02009	.01366	.12885	10839.	-.01173	.02731
17	S98T002600	.01744	-.00411	1.1242	10631.	-.00967	.04090
18	S98T002600_D	.02980	-.01122	1.5650	10487.	.01758	.04986
19	ICSA	.00229	.01279	94.546	.06189	-.00583	.00165
20	ICSAB	.45693	.00880	93.463	.21836	-.00351	.96913
21	CCV	4.8296	.00593	4.9581	5.0058	5.1006	5.1077
22	CCB	-.00056	-.00107	.00025	-.07360	-.00005	-.00088

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1308	4.9493	5.0446	5.2451	4.9770	4.9623
2	ICB	.00575	-.00023	.00031	.03396	-.00083	-.00068
3	LLS	.20504	.01980	.09896	.21031	.19747	.03412
4	ICSA	252.40	-.00538	-.01955	193.05	.00097	-.00214
5	ICSAB	249.03	.43633	-.01829	190.34	-.00033	.90336
6	PREPBLKTJA	.00523	.00134	-.00053	.07708	-.00187	Q.12554
7	S98T002598_L	1.0190	-.02105	-.50418	542.60	-1.0150	.48376
8	S98T002598	.05044	.03544	-.07554	528.13	-.13852	1.1019
9	S98T002598_D	.33465	.04229	-.03323	513.73	.01176	.78448
10	S98T002598_A	40.361	39.055	40.590	568.63	38.878	41.065
11	S98T002598_X	-.44961	-.03446	-1.2717	532.99	-.43031	-.51915
12	S98T002598_AX	4221.5	4068.5	4034.3	4571.1	3997.0	4056.7
13	CCV	5.1504	4.9519	5.0624	5.2992	5.0157	4.9807
14	CCB	-.00116	-.00025	-.00180	.02748	-.00320	-.00405
15	S98T002599	.24228	.05115	-.07526	543.36	-.05212	1.3502
16	S98T002599_D	-.10442	.04222	-.16154	543.92	-.17363	1.8813
17	S98T002600	-.02134	.08787	-.06158	561.78	-.07017	1.0629
18	S98T002600_D	.31146	.09628	-.06171	543.11	-.05733	1.2876
19	ICSA	255.82	-.00590	-.01647	197.46	-.00149	-.00987
20	ICSAB	250.79	.44035	-.01585	191.71	-.00071	.90652
21	CCV	5.1569	5.0016	5.1073	5.2385	5.0085	5.0159
22	CCB	-.00346	-.00043	-.00148	.03817	-.00297	-.00604

Analysis Report

Averages

Fri 10-30-98 06:36:45 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0345	4.9766	4.8920	5.0750	4.8018	4.9221
2	ICB	.00222	.00904	.01071	-.00162	-.00208	-.00095
3	LCS	.38640	.20785	.21263	.10146	.17728	Q.20705
4	ICSA	.02096	.03857	-.01765	.00458	.04849	-.00613
5	ICSAB	.03004	.99119	-.03831	-.00388	.06494	-.00237
6	PREPBLKTJA	-.00476	.00609	.00992	-.00372	-.00347	.00757
7	S98T002598_L	-.58463	1.0454	.43762	.10436	.46789	.35839
8	S98T002598	-.17156	.29136	.48930	-.22753	-.26185	.81936
9	S98T002598_D	-.28067	.47429	.60357	.07987	.62492	1.1751
10	S98T002598_A	40.897	40.862	40.620	40.932	39.260	41.236
11	S98T002598_X	.51473	3.9023	.50100	-1.5292	-.41828	.56311
12	S98T002598_AX	4029.6	4100.7	3953.6	4085.5	3988.2	4084.7
13	CCV	5.0576	4.9895	4.8756	5.0919	4.8326	4.9364
14	CCB	-.00923	.00856	.01533	-.00723	-.00848	-.00266
15	S98T002599	.29765	.29745	.98524	-.18263	-.05927	1.0100
16	S98T002599_D	.54965	.06891	.95514	-.08114	.44047	1.2145
17	S98T002600	8.7612	.39887	13.872	-.08095	-.10166	4.2297
18	S98T002600_D	9.2257	.31410	14.391	.14915	.03986	8.0610
19	ICSA	.02491	.01986	-.01814	.00477	.06134	-.00939
20	ICSAB	.03053	1.0059	-.01928	-.00842	.05122	-.00119
21	CCV	5.0851	5.0767	4.9283	5.1441	4.8874	4.9971
22	CCB	-.00473	.00943	.00987	-.00139	.00016	-.00442

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.9737	5.0839	.19664	5.1163	4.7998	9.8557
2	ICB	-.00627	-.00013	.00264	-.00115	-.01387	-.02814
3	LCS	.18854	.01995	.00790	.01908	.39692	.44296
4	ICSA	-.02095	.00183	.00621	.00000	.00650	-.02207
5	ICSAB	-.01830	.00182	.00444	.00069	-.02143	-.00858
6	PREPBLKTJA	-.00014	.00000	.00521	-.00063	-.01242	-.08767
7	S98T002598_L	-1.5695	-.03621	.17484	-.33330	-4.3032	-9.8294
8	S98T002598	-.26213	.00032	-.01900	-.05604	-1.1236	-4.1671
9	S98T002598_D	.28741	.00920	.12919	-.00888	-.29593	.33298
10	S98T002598_A	38.761	39.859	1.7338	40.612	38.341	73.926
11	S98T002598_X	-.42301	-.01777	-.49605	-.19130	-7.3001	-7.0339
12	S98T002598_AX	3978.8	4067.3	161.67	4044.7	4033.3	7698.8
13	CCV	5.0085	5.1222	.20279	5.1431	4.8362	9.9094
14	CCB	-.00991	-.00022	.00253	-.00162	-.01849	-.04580
15	S98T002599	-.06718	.00398	.25603	-.03592	-.94729	-3.2852
16	S98T002599_D	.14942	.00578	.06409	-.05584	-1.3269	-2.7699
17	S98T002600	-.09971	.20708	.07100	-.01768	-.57052	-2.8069
18	S98T002600_D	-.16204	.22530	.24283	.01183	-.35319	-2.9899
19	ICSA	-.02621	.00195	-.00450	-.00003	.01318	-.04469
20	ICSAB	-.01695	.00189	.00362	.00093	.01522	-.01125
21	CCV	4.9908	5.1244	.19679	5.1590	4.8983	9.8920
22	CCB	-.01102	-.00022	-.00006	-.00163	-.01562	-.05047

Analysis Report

Averages

Fri 10-30-98 06:36:45 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0094	.00670	4.7526	4.8861
2	ICB	-.00156	-.00022	-.00004	-.00080
3	LIS	.09673	-.00018	.01965	.01741
4	ICSA	.00027	.00722	-.00459	-.00483
5	ICSAB	.45897	.00737	.91785	-.00347
6	PREPBLKTJA	-.00075	-.00004	.00056	-.00084
7	S98T002598_L	-.42280	-.07616	.00531	-.26350
8	S98T002598	-.09532	-.01108	.03946	-.03395
9	S98T002598_D	.02831	.01913	.00336	.00289
10	S98T002598_A	40.147	.06498	39.178	38.563
11	S98T002598_X	-.35155	.00298	-.23684	-.08622
12	S98T002598_AX	4010.7	6.1758	3813.6	4015.7
13	CCV	5.0270	.00683	4.7648	4.9076
14	CCB	-.00260	-.00060	-.00012	-.00141
15	S98T002599	-.01671	.00139	-.02333	-.04231
16	S98T002599_D	.01971	.01403	-.00902	.01081
17	S98T002600	-.02465	-.00202	.00539	-.01587
18	S98T002600_D	-.03224	-.00194	.01190	-.04155
19	ICSA	-.00157	.00675	-.00518	-.00473
20	ICSAB	.46450	.00752	.92612	-.00351
21	CCV	5.0640	.00731	4.8433	4.9269
22	CCB	-.00312	-.00076	-.00045	-.00153

10-30-98

File #: 981030E.TXT

10/29/98 13:32
A-0004-1

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26898

Analyst: JK Scto Instrument: ICP01-1 JK 10-30-98 Book# 758783Method: LA-505-151/161 Rev/Mod C-3JK 10-30-98Ensure dose rate at 30 cm is ≤ 50 μ R/h
prior to performing this analysis.

Worklist Comment: ICP BY-105 (FUSION)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SAMPLE	S98T002601 0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01						
8	DUP	S98T002601 0 F	@ICP-F01	SOLID		
9	CCV		@ICP-QC	QC		
10	CCB		@ICP-QC	QC		
11	SERDIL	S98T002602 0 F	@ICP-F01	SOLID		
12	SAMPLE	S98T002602 0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01						

Data Entry Comments:

uploaded 11-2-98
Jim WarellValidated by:
Sailor Pang
11/2/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26898

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 DUP	S98T002602	0 F	@ICP-F01	SOLID		
14 SPK (1pp4)	S98T002602	0 F	@ICP-F01	SOLID		
15 SAMPLE	S98T002603	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested:						
			AG-F-01	AL-F-01	AS-F-01	B-F-01
			BA-F-01	BE-F-01	BI-F-01	CA-F-01
			CD-F-01	CE-F-01	CO-F-01	CR-F-01
			CU-F-01	FE-F-01	LA-F-01	LI-F-01
			MG-F-01	MN-F-01	MO-F-01	NA-F-01
			ND-F-01	NI-F-01	P-F-01	PB-F-01
			S-F-01	SB-F-01	SE-F-01	SI-F-01
			SM-F-01	SR-F-01	TI-F-01	TL-F-01
			U-F-01	V-F-01	ZN-F-01	ZR-F-01
16 DUP	S98T002603	0 F	@ICP-F01	SOLID		
17 ICSA			@ICP-QC	QC		
18 ICSAB			@ICP-QC	QC		
19 CCV			@ICP-QC	QC		
20 CCB			@ICP-QC	QC		

Final page for worklist # 26898

Analyst Signature	Date
Prepilk JTA	.25-10
5987002601	.25-10
5987002601	.25-10
5987002602	.25-10-2-8
5987002602	.25-10
5987002602	.25-10
5987002602	.25-10
5987002602	.25-10-1-9
5987002602	.25-10-1-9
5987002603	.25-10
5987002603	.25-10

Analyst Signature _____ **Date** _____

Data Entry Comments:

S = Worklist Slot Number, *R* = Replicate Number, *A* = Aliquot Code.

598T002602

$$RL = 58.181 \frac{\mu\text{g}}{\text{mL}} \times \frac{1}{2.0816 \frac{\text{g}}{\text{L}} \times \frac{1\text{L}}{1000\text{mL}}} = 2.80 \text{ } \pm 4 \frac{\mu\text{g}}{\text{g}}$$

Spike:

$$RL = \frac{\left(\frac{100.27}{41}\right) - \left(\frac{58.181}{41}\right)}{1} \times 100 = 102.7\%$$

Post Spike:

$$Na = \frac{\left(\frac{4411.6}{410}\right) - \left(\frac{293.24}{410}\right)}{10} \times 100 = 100.4\%$$

Analysis Report

Summary

Fri 10-30-98 08:06:04 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981030E	ICP2	10/30/98	18:40	DKS	Q	CONC
2	ICB	981030E	ICP2	10/30/98	18:43	DKS	Q	CONC
3	LLS	981030E	ICP2	10/30/98	18:53	DKS	Q	CONC
4	ICSA	981030E	ICP2	10/30/98	18:56	DKS	Q	CONC
5	ICSAB	981030E	ICP2	10/30/98	18:59	DKS	Q	CONC
6	PREPBLKTJA	981030E	ICP2	10/30/98	19:02	DKS	Q	CONC
7	S98T002601	981030E	ICP2	10/30/98	19:05	DKS	S	CONC
8	S98T002601_D	981030E	ICP2	10/30/98	19:08	DKS	S	CONC
9	CCV	981030E	ICP2	10/30/98	19:12	DKS	Q	CONC
10	CCB	981030E	ICP2	10/30/98	19:15	DKS	Q	CONC
11	S98T002602_L	981030E	ICP2	10/30/98	19:18	DKS	S	CONC
12	S98T002602	981030E	ICP2	10/30/98	19:21	DKS	S	CONC
13	S98T002602_D	981030E	ICP2	10/30/98	19:24	DKS	S	CONC
14	S98T002602_A	981030E	ICP2	10/30/98	19:27	DKS	S	CONC
15	S98T002602_X	981030E	ICP2	10/30/98	19:32	DKS	S	CONC
16	S98T002602_AX	981030E	ICP2	10/30/98	19:35	DKS	S	CONC
17	S98T002603	981030E	ICP2	10/30/98	19:46	DKS	S	CONC
18	S98T002603_D	981030E	ICP2	10/30/98	19:49	DKS	S	CONC
19	ICSA	981030E	ICP2	10/30/98	19:53	DKS	Q	CONC
20	ICSAB	981030E	ICP2	10/30/98	19:56	DKS	Q	CONC
21	CCV	981030E	ICP2	10/30/98	19:59	DKS	Q	CONC
22	CCB	981030E	ICP2	10/30/98	20:02	DKS	Q	CONC

10-30-98

Worklist #26898

34-105

5987602601

5987602602

5987602603

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1897 TO 1961.

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0115	4.9957	5.2069	5.0192	5.0586	5.1613
2	ICB	.00025	-.00293	-.00514	.00287	.00003	.00000
3	LLS	.02187	.09665	.20542	.09813	.09883	.01051
4	ICSA	.00316	245.74	.03384	-.01059	.00028	.00023
5	ICSAB	.94351	243.30	.05005	-.00589	.45687	.47726
6	PREPBLKTJA	-.00007	-.00080	-.00654	.00621	.00017	.00004
7	S98T002601	.13294	37.507	-.12971	.11718	.01024	-.00368
8	S98T002601_D	.03710	35.198	-.38049	.05814	.00263	-.00171
9	CCV	5.0225	4.9822	5.2102	5.0039	5.0159	5.1440
10	CCB	-.00013	-.00362	.00721	.00334	.00002	.00004
11	S98T002602_L	-.01478	58.660	.34652	.62598	1.0621	-.00413
12	S98T002602	.02723	58.181	-.14394	.11525	1.0159	-.00388
13	S98T002602_D	-.00752	57.246	-.06987	.27407	1.0472	.00352
14	S98T002602_A	37.590	100.27	42.988	40.148	40.982	41.291
15	S98T002602_X	-.44949	59.871	-2.1345	.81951	1.0414	.05311
16	S98T002602_AX	3698.1	4102.8	4182.4	4049.3	4175.4	4151.4
17	S98T002603	.08651	45.037	.18687	-.07754	1.0371	.00325
18	S98T002603_D	.02170	44.873	-.32689	.05980	1.0236	.00514
19	ICSA	.00564	247.51	.04368	-.00665	.00014	.00009
20	ICSAB	.94966	245.48	.02159	-.01253	.46096	.47856
21	CCV	5.0632	5.0436	5.2436	5.0532	5.0963	5.1820
22	CCB	-.00002	-.01242	-.00783	.00286	-.00012	-.00013

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0599	4.9945	5.1070	5.1414	5.0108	4.9704
2	ICB	.00761	.00131	-.00012	.00166	-.00148	-.00082
3	LLS	.20965	.22443	.01030	.21078	.03886	.01918
4	ICSA	.00196	255.79	.00482	.00356	-.00076	-.00277
5	ICSAB	-.00843	255.83	.93865	.00051	.46127	.46331
6	PREPBLKTJA	-.03071	.00714	.00007	-.00209	-.00030	-.00089
7	S98T002601	-.71888	.25503	.02584	.55710	.05911	.38601
8	S98T002601_D	-.99126	.14563	.00147	.03832	.03863	.31166
9	CCV	5.0096	5.0164	5.1113	5.1195	5.0113	4.9718
10	CCB	-.01322	-.00012	-.00019	-.00141	-.00025	-.00128
11	S98T002602_L	2.6159	68.712	.09186	.40434	-.00582	1.7830
12	S98T002602	4.6172	67.759	-.01501	.09567	.06505	1.8011
13	S98T002602_D	5.8585	68.281	-.02229	-.27924	.03385	1.7757
14	S98T002602_A	46.920	107.86	41.471	41.201	40.716	42.171
15	S98T002602_X	3.3291	65.735	-.16533	-2.2016	.08981	1.9068
16	S98T002602_AX	4209.4	4217.9	4120.1	4190.4	4058.6	4133.3
17	S98T002603	6.6183	44.889	-.02748	.29490	.19033	.92979
18	S98T002603_D	6.3830	45.353	-.01081	-.29357	.11175	.82848
19	ICSA	-.02191	256.49	.00518	.00156	-.00026	-.00435
20	ICSAB	.00214	256.09	.93771	.00497	.46103	.46421
21	CCV	5.0773	5.0398	5.1268	5.1912	5.0297	5.0004
22	CCB	.00253	-.00367	-.00061	-.01029	-.00001	-.00123

Analysis Report

Averages

Fri 10-30-98 08:06:04 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.8424	.00484	4.9394	5.2552	5.1112	5.1548
2	ICB	-.00045	-.00017	.00113	.14094	.00041	-.00010
3	LLS	.01748	-.00037	.09874	.52392	.02080	.02084
4	ICSA	.00275	.00931	93.948	.10530	-.00385	.00154
5	ICSAB	.45721	.00967	93.469	.00583	-.00527	.96814
6	PREPBLKTJA	.00028	.00075	.00391	Q270.25	.00090	.00044
7	S98T002601	.02567	.02101	.21779	11149.	.06771	.01372
8	S98T002601_D	.00695	-.01268	.17030	11197.	-.01725	-.00885
9	CCV	4.8175	.00526	4.9300	5.0593	5.0801	5.0969
10	CCB	-.00006	-.00065	.00050	-.03889	.00001	-.00044
11	S98T002602_L	-.17870	.08617	64.788	12364.	.07838	-.00393
12	S98T002602	-.21232	.08735	63.670	12007.	.04580	.04137
13	S98T002602_D	-.22038	.07882	66.469	12681.	-.00533	.01397
14	S98T002602_A	38.862	.08716	104.78	12665.	41.002	40.813
15	S98T002602_X	-.25055	-.38911	64.466	12608.	-.27215	-.32017
16	S98T002602_AX	3876.1	4.5694	4033.5	16652.	4176.4	4148.0
17	S98T002603	-.19933	.04297	69.603	11683.	.08403	.01378
18	S98T002603_D	-.22001	.01234	69.282	12238.	.01145	-.01337
19	ICSA	.00257	.00964	94.285	.08643	-.00431	.00198
20	ICSAB	.46221	.01026	93.645	.11980	-.00380	.98492
21	CCV	4.8815	.00544	4.9618	5.0661	5.1406	5.2268
22	CCB	.00001	-.00099	-.00015	-.07468	-.00143	-.00099

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1279	4.9598	5.0665	5.3103	5.0208	4.9800
2	ICB	.00119	-.00017	-.00261	.02977	-.00233	-.00103
3	LLS	.20936	.01977	.10222	.21489	.19741	.04345
4	ICSA	252.70	-.00550	-.01923	193.83	-.00082	-.00681
5	ICSAB	250.34	.43985	-.01700	191.57	-.00224	.91835
6	PREPBLKTJA	.00787	.00109	-.00182	.08243	-.00133	Q.12143
7	S98T002601	.56585	.05513	-.02820	498.00	-.04864	3.1957
8	S98T002601_D	.11976	.03820	-.07549	504.25	-.08019	2.6937
9	CCV	5.1244	4.9525	5.0763	5.2401	4.9771	4.9877
10	CCB	.00177	-.00010	-.00180	.03359	-.00226	-.00732
11	S98T002602_L	3.5871	.93727	-.49151	281.89	.17816	46.329
12	S98T002602	3.2245	.90374	-.05816	273.96	.28462	45.146
13	S98T002602_D	3.2524	.94827	-.05563	271.30	.28240	48.746
14	S98T002602_A	44.309	40.310	41.267	332.83	40.458	85.491
15	S98T002602_X	.86201	.79908	-1.0327	293.24	.24733	45.042
16	S98T002602_AX	4282.3	4110.2	4065.4	4411.6	4082.1	4160.1
17	S98T002603	3.0657	.93212	-.04717	277.78	.41933	51.337
18	S98T002603_D	3.0747	.89590	-.15375	284.29	.31921	44.742
19	ICSA	254.61	-.00564	-.01613	196.12	-.00386	-.01903
20	ICSAB	252.34	.44051	-.01666	194.41	-.00172	.91779
21	CCV	5.1692	4.9870	5.1237	5.3647	5.0515	5.0090
22	CCB	-.00805	-.00040	-.00263	.03934	-.00152	.00026

Analysis Report

Averages

Fri 10-30-98 08:06:04 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0046	5.0227	4.8802	5.0744	4.8044	4.9392
2	ICB	-.00739	.00133	.01400	.00051	.00222	-.00067
3	LLS	.38967	.21542	.20475	.08902	.19090	Q.20746
4	ICSA	.01994	.02752	-.01892	.01358	.04710	-.00724
5	ICSAB	.02722	.99640	-.02817	.01289	.07856	-.00189
6	PREPBLKTJA	.00874	.00492	.00403	-.00371	.00835	.03768
7	S98T002601	2.1812	.26687	1.6107	.16321	.13589	2.0810
8	S98T002601_D	1.2120	.55723	1.7372	-.23233	.05950	1.8789
9	CCV	5.0274	5.0271	4.9002	5.1086	4.8416	4.9452
10	CCB	-.01067	.01133	.00876	-.00137	-.00312	-.00094
11	S98T002602_L	39.676	2.3705	4.8195	.79340	1.8249	24.132
12	S98T002602_	40.026	1.6738	3.5740	-.00908	.39930	21.647
13	S98T002602_D	41.598	1.7624	2.5398	-.04444	.53421	20.241
14	S98T002602_A	82.384	42.349	43.800	41.453	39.084	62.807
15	S98T002602_X	35.935	4.8021	4.1555	-.1.0803	.98587	28.361
16	S98T002602_AX	4110.1	4151.8	3981.5	4117.1	4034.4	4254.1
17	S98T002603	47.558	2.2147	2.4325	.01045	-.08132	12.369
18	S98T002603_D	42.943	1.8133	2.4176	-.05288	-.20788	11.724
19	ICSA	.04449	.01624	-.03143	-.00162	.06996	-.00581
20	ICSAB	.03596	1.0031	-.03184	-.00143	.06646	-.00222
21	CCV	5.1278	5.0444	4.9542	5.1312	4.8738	4.9823
22	CCB	-.00390	.00057	.02045	-.00934	-.00086	-.00322

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	5.0137	5.1290	.19455	5.1452	4.8208	9.9198
2	ICB	-.00293	-.00008	.00290	-.00068	-.01817	-.02031
3	LLS	.19401	.02008	.00877	.01862	.39471	.45391
4	ICSA	-.02259	.00195	.00175	-.00024	.01001	-.03832
5	ICSAB	-.01787	.00194	-.00118	-.00001	-.01648	-.03251
6	PREPBLKTJA	.00321	.00018	.00248	-.00042	-.00386	-.06510
7	S98T002601	.14092	.03716	.35657	-.001630	-.59942	-.2.6841
8	S98T002601_D	-.15714	.03172	.09423	-.04602	-.42072	-3.7995
9	CCV	4.9819	5.0963	.20420	5.1403	4.9043	9.8387
10	CCB	-.00698	-.00013	-.00373	-.00094	-.02028	-.02951
11	S98T002602_L	.43479	48.091	.86389	.15896	-3.5444	72.940
12	S98T002602_	.23967	46.993	1.3055	.27611	-.60373	72.326
13	S98T002602_D	.10706	48.198	1.2429	.26387	-.17643	76.595
14	S98T002602_A	39.529	89.873	2.6590	41.875	38.539	153.46
15	S98T002602_X	-3.8709	48.711	-.10605	-.03680	-4.7632	59.998
16	S98T002602_AX	4060.5	4191.1	161.99	4096.2	4047.5	7912.7
17	S98T002603	-.05491	49.130	1.4745	.17616	-.63053	80.061
18	S98T002603_D	-.37088	49.224	1.2983	.16542	-.80208	80.033
19	ICSA	-.02169	.00192	-.00042	-.00025	.00101	-.02558
20	ICSAB	-.01918	.00193	-.00016	.00068	-.02170	-.01132
21	CCV	5.0550	5.1652	.19854	5.1877	4.8453	9.9991
22	CCB	-.01095	-.00017	-.00398	-.00094	-.02344	-.05114

Analysis Report

Averages

Fri 10-30-98 08:06:04 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0368	.00683	4.7652	4.9184
2	ICB	-.00140	-.00006	-.00014	-.00019
3	LLS	.09867	-.00003	.01982	.01765
4	ICSA	.00012	.00722	-.00449	-.00485
5	ICSAB	.46387	.00745	.92402	-.00345
6	PREPBLKTJA	.00082	.00041	.00030	.00050
7	S98T002601	.01854	.01399	-.02329	-.01131
8	S98T002601_D	-.03738	-.00481	-.00961	-.02969
9	CCV	5.0258	.00694	4.7862	4.9006
10	CCB	-.00154	-.00030	-.00042	-.00064
11	S98T002602_L	-.07424	.06540	.18661	-.06404
12	S98T002602	.04291	.04638	.25688	.02145
13	S98T002602_D	.02750	.04135	.24628	.02261
14	S98T002602_A	40.740	.07295	39.505	39.483
15	S98T002602_X	-1.0019	-.18617	-.09875	-.43391
16	S98T002602_AX	4048.7	6.1813	3822.6	4063.0
17	S98T002603	.01319	.03024	.25451	-.01013
18	S98T002603_D	-.08152	.01740	.26582	-.04882
19	ICSA	.00009	.00722	-.00469	-.00441
20	ICSAB	.46546	.00745	.92774	-.00356
21	CCV	5.0713	.00703	4.8091	4.9454
22	CCB	-.00329	-.00060	-.00063	-.00124

JK

10-30-98

LABCORE Data Entry Template for Worklist# 26947

Analyst:

*JK Seto*Instrument: ICPOT *11/10/98*Book# *77B30*Method: LA-505-151/161 Rev/Mod *C-3*

Worklist Comment: ICP BY-105 (FUSION)

*Ensure dose rate of 30 c/h is $\leq$ 50 nrem/hr
prior to performing this analysis.*

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SAMPLE	S98T002845	0 F	@ICP-F01	SOLID	98000437 BY-105 (2)
	Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01					
8	DUP	S98T002845	0 F	@ICP-F01	SOLID	
9	SAMPLE	S98T002918	0 F	@ICP-F01	SOLID	98000490 BY-105 (2)
	Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01					
10	DUP	S98T002918	0 F	@ICP-F01	SOLID	
11	CCV		@ICP-QC	QC		
12	CCB		@ICP-QC	QC		

Data Entry Comments:

*uploaded 12-14-98**John W. Wrenell**26947 12-14-98 JW*

Validated by:

*Saul R. Pung**12-14-98*

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26947

S Type	Sample#	R A	Test	Matrix	Group#	Project
13 SERDIL	S98T002919	0 F	@ICP-F01	SOLID		
14 SAMPLE	S98T002919	0 F	@ICP-F01	SOLID	98000490	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
15 DUP	S98T002919	0 F	@ICP-F01	SOLID		
16 SPK (1ppb)	S98T002919	0 F	@ICP-F01	SOLID		
17 ICSA	SPK (10ppm) 11-10-98		@ICP-QC	QC		
18 ICSAB			@ICP-QC	QC		
19 CCV			@ICP-QC	QC		
20 CCB			@ICP-QC	QC		

Final page for worklist # 26947

Analyst Signature   **Date**
Prepbk TIR , 12-10-98 , JF 1
5987001845 , 12-10-98 , JF 41
5987001845-D , 12-10-98 , 41
5987001845 , 12-10-98 , JF 41
5987001845-D , 12-10-98 , 41

Analyst Signature   **Date**
5987001845-L , 12-10-98 , JF 105
5987001845 , 12-10-98 , 41
5987001845-D , 12-10-98 , 41
5987001845-Q , 12-10-98 , 41
5987001845-X , 12-10-98 , 410
5987001845-QX , 12-10-98 , 410

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

Summary

Mon 12-14-98 11:30:52 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981214B	ICP2	12/14/98	10:07	DKS	Q	CONC
2	ICB	981214B	ICP2	12/14/98	10:10	DKS	Q	CONC
3	LLS	981214B	ICP2	12/14/98	10:13	DKS	Q	CONC
4	ICSA	981214B	ICP2	12/14/98	10:16	DKS	Q	CONC
5	ICSAB	981214B	ICP2	12/14/98	10:19	DKS	Q	CONC
6	PREPBLKTJA	981214B	ICP2	12/14/98	10:22	DKS	Q	CONC
7	S98T002845	981214B	ICP2	12/14/98	10:25	DKS	S	CONC
8	S98T002845_D	981214B	ICP2	12/14/98	10:28	DKS	S	CONC
9	S98T002918	981214B	ICP2	12/14/98	10:32	DKS	S	CONC
10	S98T002918_D	981214B	ICP2	12/14/98	10:35	DKS	S	CONC
11	CCV	981214B	ICP2	12/14/98	10:38	DKS	Q	CONC
12	CCB	981214B	ICP2	12/14/98	10:43	DKS	Q	CONC
13	S98T002919_L	981214B	ICP2	12/14/98	10:47	DKS	S	CONC
14	S98T002919	981214B	ICP2	12/14/98	10:50	DKS	S	CONC
15	S98T002919_D	981214B	ICP2	12/14/98	10:53	DKS	S	CONC
16	S98T002919_A	981214B	ICP2	12/14/98	10:56	DKS	S	CONC
17	S98T002919_X	981214B	ICP2	12/14/98	11:00	DKS	S	CONC
18	S98T002919_AX	981214B	ICP2	12/14/98	11:03	DKS	S	CONC
19	ICSA	981214B	ICP2	12/14/98	11:18	DKS	Q	CONC
20	ICSAB	981214B	ICP2	12/14/98	11:21	DKS	Q	CONC
21	CCV	981214B	ICP2	12/14/98	11:24	DKS	Q	CONC
22	CCB	981214B	ICP2	12/14/98	11:27	DKS	Q	CONC

M L&D

12-14-98

Worklist # 26947

B 4-105

5987002845

5987002918

5987002919

5987002919

$$AL = 3.5953 \frac{\mu g}{ml} \times \frac{1}{1.9960 \frac{g}{L} \times \frac{1L}{100ml}} = 1.80 \mu g + 3 \frac{\mu g}{g}$$

Post Spike:

$$AL = \left(\frac{42.374}{41} \right) - \left(\frac{3.5953}{41} \right) \times 100 = 94.6\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1904 TO 1908.

Analysis Report

Averages

Mon 12-14-98 11:30:52 AM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9776	4.8207	5.0229	5.1537	4.8300	5.0462
2	ICB	-.00049	-.00505	.00062	.00137	.00009	.00000
3	LLS	.02075	.09790	.19897	.10062	.09702	.01030
4	ICSA	.00075	238.42	.05474	-.00786	.00023	.00017
5	ICSAB	.95023	240.17	.03836	-.00511	.45668	.47342
6	PREPBLKTJA	-.00145	-.00264	-.00082	-.00230	.00006	.00004
7	S98T002845	-.01343	35.300	-.06119	-.11667	.35665	.00206
8	S98T002845_D	-.01976	36.302	.04574	.03465	.37899	.00020
9	S98T002918	.00652	80.467	.15693	.01878	.00711	-.00328
10	S98T002918_D	-.04800	56.474	-.23767	-.07562	.00245	-.00501
11	CCV	4.9899	4.8250	5.0296	5.1408	4.8109	5.0834
12	CCB	-.00061	.00151	-.00471	-.00091	-.00002	.00012
13	S98T002919_L	-.32101	3.0935	-1.4803	-1.0355	-.01355	.00047
14	S98T002919	-.05544	3.5953	-.16546	-.09458	-.00126	.00202
15	S98T002919_D	-.06062	3.2217	-.18741	.05628	.00045	-.00328
16	S98T002919_A	37.077	42.374	41.003	40.964	37.098	39.884
17	S98T002919_X	-.51594	3.9525	.36846	-1.8790	.02156	.05297
18	S98T002919_AX	3527.9	3833.6	4014.1	3969.7	3896.1	3902.2
19	ICSA	.00403	234.00	.05315	-.00604	.00025	.00030
20	ICSAB	.94936	237.65	.04118	-.00719	.44986	.46456
21	CCV	5.0100	4.8264	5.0155	5.1150	4.8258	5.0358
22	CCB	.00050	-.00326	-.00847	.00275	.00003	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0222	5.1509	5.0844	4.9975	4.9638	4.9814
2	ICB	-.00195	.00807	-.00121	-.00190	-.00083	-.00127
3	LLS	.19063	.23077	.00867	.19429	.03943	.01881
4	ICSA	-.02346	251.63	.00369	-.00653	-.00182	-.00475
5	ICSAB	-.01393	256.22	.94321	-.00370	.46128	.46699
6	PREPBLKTJA	.01895	.00931	-.00147	-.00351	-.00029	-.00161
7	S98T002845	1.9137	21.961	-.04176	-.22421	.01317	2.1698
8	S98T002845_D	1.8619	22.724	-.02538	-.21082	-.01748	1.9185
9	S98T002918	-.36257	.69668	-.05104	-.25998	-.04746	1.2250
10	S98T002918_D	-.25631	.68431	-.05367	-.54827	-.08218	.77829
11	CCV	5.0630	5.1914	5.1158	4.9936	5.0133	5.0294
12	CCB	-.01520	.00953	-.00043	-.00126	-.00095	-.00050
13	S98T002919_L	-.58953	1.8171	-.27509	-1.8182	-.32134	-.04227
14	S98T002919	-.21452	.92948	-.05650	-.50739	-.01736	.31592
15	S98T002919_D	-.1.8678	.49697	-.06272	-.58430	-.05265	.29087
16	S98T002919_A	39.601	43.168	41.409	39.334	40.310	40.946
17	S98T002919_X	-4.2018	3.4342	-.19355	-.75007	-.53165	-.08179
18	S98T002919_AX	4063.2	4155.6	4017.6	3921.7	4031.8	4053.1
19	ICSA	-.02073	255.84	.00337	.00390	-.00180	-.00277
20	ICSAB	.00930	256.13	.93649	-.00652	.45699	.46337
21	CCV	5.0122	5.2188	5.1077	4.9907	5.0101	5.0219
22	CCB	-.00483	.00998	-.00139	-.00600	-.00035	-.00052

Analysis Report

Averages

Mon 12-14-98 11:30:52 AM

page 3

#	Sample Name	Cu	Su	Fe	K	La	Li
1	ICV	4.7263	.00911	4.7835	4.6697	5.0180	4.9818
2	ICB	-.00105	.00019	-.00029	-.06957	.00022	-.00020
3	LLS	.01685	.00035	.09679	.38679	.09958	.01993
4	ICSA	.00163	.03351	91.488	-.09691	-.00484	.00191
5	ICSAB	.45984	.03202	92.633	-.19594	-.00440	.97464
6	PREPBLKTJA	-.00013	.00082	.00229	Q257.86	-.00030	.00040
7	S98T002845	-.09432	.07334	26.377	11903.	.00165	.03148
8	S98T002845_D	-.08226	.06179	27.477	11381.	.00600	.01488
9	S98T002918	-.00339	.04106	.63140	11402.	.00046	.01250
10	S98T002918_D	-.01077	.00314	.51480	11442.	-.05468	.00432
11	CCV	4.7055	.00945	4.8123	4.8097	5.0134	4.9255
12	CCB	-.00090	.00055	.00031	-.16001	.00024	.00010
13	S98T002919_L	-.04645	.00991	.46787	11610.	-.21450	.00018
14	S98T002919	.01218	.02002	.16232	11218.	-.05793	.00015
15	S98T002919_D	-.01981	.04036	.15277	11197.	-.06269	.02077
16	S98T002919_A	37.016	.13442	38.739	11258.	39.204	37.761
17	S98T002919_X	-.34855	.68109	-.50824	11255.	-.12330	.49509
18	S98T002919_AX	3657.4	7.7886	3832.5	15107.	3897.3	3819.8
19	ICSA	.00211	.02026	91.387	-.04203	-.00397	.00343
20	ICSAB	.45072	.02708	91.904	-.04566	-.00505	.95226
21	CCV	4.7039	.01028	4.8031	4.6884	5.0062	4.9890
22	CCB	-.00050	.00036	-.00007	-.22562	.00026	.00010

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1203	4.9090	5.0353	4.9147	4.8970	4.9917
2	ICB	-.00242	.00000	.00205	-.00241	-.00257	.00055
3	LLS	.19846	.02012	.10322	.18598	.19581	.04096
4	ICSA	245.89	-.00513	-.01692	189.77	-.00292	-.00824
5	ICSAB	247.18	.44281	-.01746	189.45	-.00260	.92356
6	PREPBLKTJA	-.00081	.00191	-.00113	.04633	-.00274	Q.12668
7	S98T002845	1.0396	.44122	-.07370	356.12	.12413	28.873
8	S98T002845_D	1.2825	.43475	.05649	355.29	.12151	32.048
9	S98T002918	-.00303	.07758	.03768	422.81	-.04642	8.0896
10	S98T002918_D	-.30001	.06451	-.04698	462.64	-.21105	7.2355
11	CCV	5.1018	4.9350	5.0777	4.8674	4.8787	5.0015
12	CCB	-.00210	.00005	.00158	.00172	-.00200	.00056
13	S98T002919_L	-1.5241	.05524	-.16326	514.55	-.36994	5.1054
14	S98T002919	-.38397	.05360	.06955	497.11	-.03521	5.3325
15	S98T002919_D	-.37192	.06639	-.04622	490.52	-.17005	8.8571
16	S98T002919_A	40.444	39.056	41.046	534.26	37.970	45.831
17	S98T002919_X	-2.1970	.21965	.01009	488.04	-.45142	3.2179
18	S98T002919_AX	4074.4	4018.9	4014.7	4302.0	3790.4	4039.4
19	ICSA	241.84	-.00424	-.01534	182.30	-.00212	-.00816
20	ICSAB	244.93	.43953	-.01493	185.99	-.00226	.91356
21	CCV	5.1118	4.9437	5.0841	4.9216	4.8830	4.9826
22	CCB	-.00643	-.00005	.00284	-.00033	-.00162	-.00336

Analysis Report

Averages

Mon 12-14-98 11:30:52 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	4.9947	5.0231	4.9205	5.0677	4.8201	5.1993
2	ICB	-.00292	-.01643	.01794	-.00846	.00140	.00485
3	LLS	.37922	.20029	.19381	.08976	.20131	Q.24643
4	ICSA	.02770	.01144	-.02616	.00205	.06502	-.00008
5	ICSAB	.02295	.98155	-.02812	-.00842	.06757	.00087
6	PREPBLKTJA	-.00236	-.00847	.00238	.00416	.01489	.01821
7	S98T002845	18.504	.24605	2.4814	-.13766	.81909	6.2943
8	S98T002845_D	17.200	.63039	2.9456	-.32081	.58476	6.1326
9	S98T002918	4.9344	-.06914	1.6462	-.24701	.70831	3.2866
10	S98T002918_D	3.7182	-.24812	1.8790	-.23881	.43631	2.8516
11	CCV	5.0404	5.0670	4.8944	5.0992	4.8858	5.2176
12	CCB	-.00155	-.00661	.01491	-.00230	.01583	.00355
13	S98T002919_L	2.2076	-3.6569	2.4177	-1.2937	-.00038	3.8073
14	S98T002919_X	.66481	-.37972	1.8726	-.36932	.62898	1.2330
15	S98T002919_D	.82084	-.90236	2.0904	-.36237	.58475	1.0327
16	S98T002919_A	41.298	40.970	42.081	41.117	40.046	44.222
17	S98T002919_AX	3.1447	-3.7908	2.1677	-1.2286	6.5776	8.6568
18	S98T002919_AX	3992.3	4115.2	4004.8	4048.4	3899.0	4009.2
19	ICSA	.04632	.00818	-.01474	.00001	.07553	.00743
20	ICSAB	.04396	.99388	-.02980	.00382	.06467	.00424
21	CCV	5.0691	5.0937	4.9371	5.0937	4.8473	5.2243
22	CCB	.00098	-.00351	.00583	-.00644	.02037	.00443

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8614	4.8964	.24013	5.0897	4.7693	9.6739
2	ICB	.00037	.00000	-.00214	-.00066	.00239	-.00041
3	LLS	.19321	.01977	.00715	.01995	.38991	.46212
4	ICSA	-.01934	.00175	.00700	.00090	-.00644	-.05832
5	ICSAB	-.02026	.00167	.00453	.00067	-.01373	-.06663
6	PREPBLKTJA	.00297	.00008	-.00250	-.00041	-.00714	-.05278
7	S98T002845	.36750	16.269	.61987	.10432	-.73342	29.305
8	S98T002845_D	.19410	17.132	.56895	.09432	.17298	30.470
9	S98T002918	.31852	.03748	.02582	-.01631	-.09692	-2.0155
10	S98T002918_D	-.13941	.03252	-.09703	.01019	-.33250	-3.5440
11	CCV	4.8595	4.8838	.24855	5.0911	4.8189	9.6286
12	CCB	.00172	.00004	-.00292	-.00023	-.01434	.00714
13	S98T002919_L	-.18279	.01720	-1.0644	-.09210	-3.4188	-3.4583
14	S98T002919_X	.04157	.01888	-.10499	-.00781	-.38821	-2.8979
15	S98T002919_D	.30922	.03570	-.33185	-.03561	.13058	-1.9997
16	S98T002919_A	38.140	38.125	2.0474	40.466	37.582	73.298
17	S98T002919_X	5.6128	.08318	-1.8684	-.18693	-6.5937	17.467
18	S98T002919_AX	3750.5	3857.9	193.14	3922.8	3963.6	7331.8
19	ICSA	-.00559	.00156	.00731	.00070	-.01542	-.05321
20	ICSAB	-.01160	.00167	.00256	.00024	-.00291	-.05903
21	CCV	4.8416	4.8787	.24017	5.0795	4.8167	9.6153
22	CCB	.00111	.00000	-.00647	-.00023	-.00065	-.00643

Analysis Report

Averages

Mon 12-14-98 11:30:52 AM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9922	.00811	4.7515	4.8593
2	ICB	.00034	-.00007	.00000	.00016
3	LLS	.09868	.00019	.01981	.01897
4	ICSA	-.00007	.00739	-.00348	-.00420
5	ICSAH	.46637	.00755	.93452	-.00365
6	PREPBLKTJA	.00179	.00002	.00004	.00142
7	S98T002845	.09295	.03253	.07982	.06257
8	S98T002845_D	.06272	.02239	.07949	.05313
9	S98T002918	.08683	.02334	-.00353	.07310
10	S98T002918_D	-.00038	-.00492	-.01970	.01055
11	CCV	5.0095	.00829	4.7909	4.8695
12	CCB	.00117	.00000	-.00018	.00110
13	S98T002919_L	.14828	-.03066	.00049	.14413
14	S98T002919	.03319	-.00198	-.01871	.03471
15	S98T002919_D	.08101	.01370	-.01868	.07532
16	S98T002919_A	40.169	.09416	39.502	38.466
17	S98T002919_X	1.3649	.34602	.09088	1.1006
18	S98T002919_AX	3956.6	7.1342	3878.1	3907.3
19	ICSA	.00464	.00840	-.00326	-.00188
20	ICSAH	.46206	.00809	.93509	-.00228
21	CCV	5.0058	.00845	4.7981	4.8602
22	CCB	.00036	-.00007	.00010	.00078

DK AD

12-14-98

LABCORE Data Entry Template for Worklist# 27097

Analyst: JSa70 Instrument: ICP01 2 11-11-98 Book# 77830Method: LA-505-451/161 Rev/Mod C-3Worklist Comment: ICP BY-105 (FUSION)Ensure dose rate of 30 c/h is 2.50 urem/hr
prior to performing this analysis.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SERDIL	S98T002844	0 F	@ICP-F01	SOLID	
8	SAMPLE	S98T002844	0 F	@ICP-F01	SOLID	98000437 BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
9	DUP	S98T002844	0 F	@ICP-F01	SOLID	
10	SPK (1 ppm)	S98T002844	0 F	@ICP-F01	SOLID	
11	CCV	S98T002844	0 F	@ICP-QC	QC	
12	CCB			@ICP-QC	QC	
13	SERDIL	S98T002739	0 F	@ICP-F01	SOLID	
14	SAMPLE	S98T002739	0 F	@ICP-F01	SOLID	98000490 BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 ,						

Data Entry Comments:

upgraded 12-28-98
John WarrallValidated by:
Sally M. Ray
12/28/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27097

S Type	Sample#	R A	Test	Matrix	Group#	Project
			MN-F-01 ,	MO-F-01 ,	NA-F-01 ,	ND-F-01 , NI-F-01 , P-F-01 ,
			PB-F-01 ,	S-F-01 ,	SB-F-01 ,	SE-F-01 , SI-F-01 , SM-F-01 ,
			SR-F-01 ,	TI-F-01 ,	TL-F-01 ,	U-F-01 , V-F-01 , ZN-F-01 ,
			ZR-F-01			
15 DUP	S98T002739	0 F	@ICP-F01	SOLID		
16 SPK (1ppm)	S98T002739	0 F	@ICP-F01	SOLID		
17 SAMPLE	S98T002712	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested:						
			AG-F-01 ,	AL-F-01 ,	AS-F-01 ,	B-F-01 , BA-F-01 ,
			BE-F-01 ,	BI-F-01 ,	CA-F-01 ,	CD-F-01 , CE-F-01 , CO-F-01 ,
			CR-F-01 ,	CU-F-01 ,	FE-F-01 ,	LA-F-01 , LI-F-01 , MG-F-01 ,
			MN-F-01 ,	MO-F-01 ,	NA-F-01 ,	ND-F-01 , NI-F-01 , P-F-01 ,
			PB-F-01 ,	S-F-01 ,	SB-F-01 ,	SE-F-01 , SI-F-01 , SM-F-01 ,
			SR-F-01 ,	TI-F-01 ,	TL-F-01 ,	U-F-01 , V-F-01 , ZN-F-01 ,
			ZR-F-01			
18 DUP	S98T002712	0 F	@ICP-F01	SOLID		
19 ICSA			@ICP-QC	QC		
20 ICSAB			@ICP-QC	QC		
21 CCV			@ICP-QC	QC		
22 CCB			@ICP-QC	QC		

Final page for worklist # 27097

DK *[Signature]* 12-21-98

Analyst Signature	Date
Prep6117JR	12-15-10
S987001844-L	12-15-10-2-8
S987001844	12-15-10
S987001844-D	12-15-10
S987001844-Q	12-15-10
S987001844-X	12-15-10-1-9
S987001844-QX	12-15-10-1-9

Analyst Signature	Date
S987002739-L	12-15-10-2-8
S987002739	12-15-10
S987002739-D	12-15-10
S987002739-Q	12-15-10
S987002739-X	12-15-10-1-9
S987002739-QX	12-15-10-1-9
S987002712	12-15-10
S987002712-D	12-15-10

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot-Code.

S98T002844

HNF-1663 REV. 0

$$RL = 27.413 \frac{\mu g}{ml} \times \frac{1}{1.9852 \frac{g}{L} \times \frac{1L}{1000ml}} = 1.3824 \frac{\mu g}{g}$$

Post Spike:

$$RL = \frac{\left(\frac{68.263}{41} \right) - \left(\frac{27.413}{41} \right)}{1} \times 100 = 99.6\%$$

$$Na = \frac{\left(\frac{4552.6}{410} \right) - \left(\frac{451.51}{410} \right)}{10} \times 100 = 100.0\%$$

S98T002739

$$RL = 105.64 \frac{\mu g}{ml} \times \frac{1}{2.0098 \frac{g}{L} \times \frac{1L}{1000ml}} = 5.2724 \frac{\mu g}{g}$$

Post Spike:

$$RL = \frac{\left(\frac{136.62}{41} \right) - \left(\frac{105.64}{41} \right)}{1} \times 100 = 75.6\%$$

$$Na = \frac{\left(\frac{4335.9}{410} \right) - \left(\frac{431.00}{410} \right)}{10} \times 100 = 95.2\%$$

$$RL = \frac{\left(\frac{4003.0}{410} \right) - \left(\frac{107.01}{410} \right)}{10} \times 100 = 95.0\%$$

Analysis Report

Summary

Mon 12-21-98 12:43:06 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981221A	ICP2	12/21/98	09:15	DKS	Q	CONC
2	ICB	981221A	ICP2	12/21/98	09:18	DKS	Q	CONC
3	LLS	981221A	ICP2	12/21/98	09:21	DKS	Q	CONC
4	ICSA	981221A	ICP2	12/21/98	09:24	DKS	Q	CONC
5	ICSAB	981221A	ICP2	12/21/98	09:27	DKS	Q	CONC
6	PREPBLKTJA	981221A	ICP2	12/21/98	09:30	DKS	Q	CONC
7	S98T002844_L	981221A	ICP2	12/21/98	09:33	DKS	S	CONC
8	S98T002844	981221A	ICP2	12/21/98	09:37	DKS	S	CONC
9	S98T002844_D	981221A	ICP2	12/21/98	09:41	DKS	S	CONC
10	S98T002844_A	981221A	ICP2	12/21/98	10:02	DKS	S	CONC
11	S98T002844_X	981221A	ICP2	12/21/98	10:07	DKS	S	CONC
12	S98T002844_AX	981221A	ICP2	12/21/98	10:10	DKS	S	CONC
13	CCV	981221A	ICP2	12/21/98	11:30	DKS	Q	CONC
14	CCB	981221A	ICP2	12/21/98	11:34	DKS	Q	CONC
15	S98T002739_L	981221A	ICP2	12/21/98	11:37	DKS	S	CONC
16	S98T002739	981221A	ICP2	12/21/98	11:40	DKS	S	CONC
17	S98T002739_D	981221A	ICP2	12/21/98	11:44	DKS	S	CONC
18	S98T002739_A	981221A	ICP2	12/21/98	11:47	DKS	S	CONC
19	S98T002739_X	981221A	ICP2	12/21/98	11:51	DKS	S	CONC
20	S98T002739_AX	981221A	ICP2	12/21/98	11:55	DKS	S	CONC
21	S98T002712	981221A	ICP2	12/21/98	12:12	DKS	S	CONC
22	S98T002712_D	981221A	ICP2	12/21/98	12:16	DKS	S	CONC
23	ICSAB	981221A	ICP2	12/21/98	12:20	DKS	Q	CONC
24	ICSA	981221A	ICP2	12/21/98	12:23	DKS	Q	CONC
25	ICSAB	981221A	ICP2	12/21/98	12:26	DKS	Q	CONC
26	CCV	981221A	ICP2	12/21/98	12:29	DKS	Q	CONC
27	CCB	981221A	ICP2	12/21/98	12:40	DKS	Q	CONC

12/21/98

JK 168

12-21-98

Work/157 # 27097

BY-105

598T002844

598T002739

598T002712

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1912 TO 1916.

Analysis Report

Averages

Mon 12-21-98 12:43:06 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0163	4.9627	5.1156	5.2325	4.9477	5.1679
2	ICB	.00000	.00069	-.00197	.00370	.00000	-.00004
3	LLS	.02117	.10859	.20994	.10013	.09934	.01055
4	ICSA	.00275	243.76	-.02682	-.00414	.00025	.00012
5	ICSAB	.95671	245.45	-.01771	.00001	.46542	.47817
6	PREPBLKTJA	-.00005	.00886	-.00501	.00045	.00005	-.00008
7	S98T002844_L	-.04214	28.225	-.31312	-.18782	.05059	-.00013
8	S98T002844_A	.01016	27.413	-.19568	-.01701	.04921	.00348
9	S98T002844_D	-.01393	32.523	-.17289	-.13085	.05243	.00348
10	S98T002844_X	37.493	68.263	41.311	41.256	38.870	40.990
11	S98T002844_X	.28345	28.240	-.62922	.76030	.05585	-.05357
12	S98T002844_AX	3459.3	4041.7	4145.2	4080.2	4145.5	4074.2
13	CCV	4.9679	4.8676	5.0538	5.0690	4.8313	5.0552
14	CCB	-.00110	-.00600	-.00205	-.00092	-.00013	-.00004
15	S98T002739_L	-.19214	106.05	-.18680	-.09490	-.00520	.00041
16	S98T002739_A	.00736	105.64	.03861	-.15207	.00995	.00023
17	S98T002739_D	.03332	105.14	-.08974	-.07584	.01041	-.00155
18	S98T002739_X	33.392	136.62	38.823	37.779	35.449	38.125
19	S98T002739_AX	-.26380	107.01	3.6825	-1.7092	.00463	.01811
20	S98T002739_AX	2810.7	4003.0	4055.2	3951.6	3971.5	3956.2
21	S98T002712	.12358	39.472	-.04711	.00609	.09747	-.00204
22	S98T002712_D	.01854	32.174	-.08548	-.05175	.08595	-.00185
23	ICSAB	0.03313	242.05	0.01624	0.00913	0.00018	0.00034
24	ICSA	.02468	242.18	.01713	-.00255	.00022	.00034
25	ICSAB	.96759	244.66	.01268	-.00468	.46693	.47787
26	CCV	5.0218	4.9108	5.1072	5.1509	4.9081	5.1291
27	CCB	.00785	.00238	-.00263	-.00231	.00000	.00004

12-23-98

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0680	5.2041	5.1703	5.1239	5.0187	5.0551
2	ICB	.00617	-.00180	-.00027	-.00161	.00116	.00041
3	LLS	.18550	.23471	.01039	.20841	.04243	.02158
4	ICSA	-.04610	256.42	.00458	.01767	.00081	-.00149
5	ICSAB	-.00284	255.59	.95141	.00388	.46346	.46871
6	PREPBLKTJA	-.02334	.00929	.00062	-.00161	.00214	.00124
7	S98T002844_L	-2.2712	4.5256	-.10918	.14338	.14147	1.7406
8	S98T002844_A	-.22378	6.5923	-.00359	-.06189	-.00810	1.6995
9	S98T002844_D	.15741	4.1576	-.00771	-.01980	.04492	1.8634
10	S98T002844_X	40.395	45.922	41.768	40.995	40.347	42.356
11	S98T002844_X	-7.4210	3.0852	.04397	-.25791	.66489	2.1402
12	S98T002844_AX	4163.2	4115.3	4089.5	4153.2	4032.4	4098.2
13	CCV	5.0593	5.1400	5.1036	5.0010	4.9560	4.9973
14	CCB	-.01668	-.00372	-.00026	-.01240	-.00081	-.00033
15	S98T002739_L	-2.3282	.41500	-.08218	-2.0760	-.02707	.69054
16	S98T002739_A	-.20768	1.1719	.01573	-.22909	.04357	.78006
17	S98T002739_D	-.57563	.48211	-.00143	-.22667	.09573	.82933
18	S98T002739_X	38.131	40.160	39.541	35.082	38.688	39.818
19	S98T002739_AX	-10.284	-.62285	-.06660	-1.7306	.28667	.40964
20	S98T002739_AX	4095.4	4068.7	4020.1	3983.7	3981.2	4032.7
21	S98T002712	.23417	24.937	.04976	.04832	.07593	5.8196
22	S98T002712_D	.03996	23.753	.00209	-.37031	-.03403	4.6449

Analysis Report

Averages

Mon 12-21-98 12:43:06 PM

page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
23	ICSA	.00065	252.70	Q.00304	.00304	Q.00011	Q.00403
24	ICSA	.00859	255.08	.00366	.00518	.00127	-.00041
25	ICSA	-.02133	254.72	.94569	.00282	.46066	.47030
26	CCV	5.1397	5.2229	5.1709	5.0683	5.0253	5.0581
27	CCB	-.01075	-.00275	.00011	-.00584	-.00150	-.00073

12-21-98

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.8111	.00841	4.8940	5.1531	5.1361	5.1588
2	ICB	-.00005	.00032	.00035	-.08360	.00017	.00032
3	LLS	.01804	.00012	.09977	.41406	.10339	.02156
4	ICSA	.00027	.02881	93.790	.07615	-.00222	.00314
5	ICSA	.46390	.03409	93.638	-.06494	-.00408	1.0190
6	PREPBLKTJA	.00047	.00145	.00218	Q274.48	.00029	.00076
7	S98T002844_L	.00529	.02033	3.5312	11423.	.09701	.04415
8	S98T002844_X	.03764	.07562	3.5434	11037.	.00169	.06653
9	S98T002844_D	.03197	.10426	3.7954	10988.	.00483	.08872
10	S98T002844_A	38.565	.06389	42.803	11634.	41.014	40.008
11	S98T002844_X	.35073	.02077	3.7374	11633.	.10702	-.00007
12	S98T002844_AX	3850.4	6.4383	3933.2	15762.	4136.2	4121.9
13	CCV	4.6922	.00825	4.8012	5.1407	5.0260	4.9456
14	CCB	-.00074	.00037	-.00153	-.08260	-.00161	.00010
15	S98T002739_L	.01363	.00124	1.0483	11851.	-.25387	.04458
16	S98T002739_X	.02763	.01511	1.2301	11653.	.01973	.01793
17	S98T002739_D	.03582	.04480	1.5320	11435.	.00366	.04010
18	S98T002739_A	35.208	.20221	37.874	11268.	36.959	36.032
19	S98T002739_X	-.25735	.21499	.92704	11744.	.07209	.22215
20	S98T002739_AX	3699.4	6.8578	3843.7	15441.	3968.2	3886.2
21	S98T002712_X	-.01006	.08119	13.532	11006.	.01745	.06191
22	S98T002712_D	-.02580	.05522	11.653	12037.	-.05341	.03533
23	ICSA	Q.00186	.03087	92.789	-.08264	-.00489	Q.00281
24	ICSA	.00297	.03092	93.045	.08872	-.00361	.00325
25	ICSA	.46250	.03664	93.413	-.11408	-.00469	.99891
26	CCV	4.7470	.00933	4.8459	5.1139	5.0873	5.0053
27	CCB	-.00035	.00056	-.00026	-.06167	-.00036	.00065

12-21-98

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.2543	4.9991	5.1083	5.1288	5.0081	5.0122
2	ICB	.00061	.00005	.00000	.01286	.00027	.00185
3	LLS	.21627	.02016	.10267	.21826	.20502	.04077
4	ICSA	252.57	-.00540	-.01209	194.02	.00312	-.00382
5	ICSA	252.93	.44483	-.01494	197.16	.00051	.93275
6	PREPBLKTJA	.00177	.00147	.00214	.09221	.00330	Q.16515
7	S98T002844_L	1.4378	.14316	.09880	437.87	.08697	10.939
8	S98T002844_X	.29918	.13729	.04004	420.15	.08727	10.895
9	S98T002844_D	.31436	.17316	.00848	423.47	.14790	10.747
10	S98T002844_A	42.267	39.403	41.422	488.50	39.924	51.825
11	S98T002844_X	1.2926	.05731	.31706	451.51	1.1805	10.917
12	S98T002844_AX	4241.0	4093.4	4037.1	4552.6	4032.0	4099.2
13	CCV	5.1345	4.9034	5.0444	4.9196	4.8833	4.9695
14	CCB	-.00757	-.00002	-.00108	.00993	-.00173	.00130

Analysis Report

Averages

Mon 12-21-98 12:43:06 PM

page 4

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
15	S98T002739_L	-.07583	.08625	.03132	441.78	-.34273	9.7521
16	S98T002739	.07150	.11096	.01758	431.74	-.00514	9.3955
17	S98T002739_D	.20465	.11281	.05625	435.32	.10830	11.584
18	S98T002739_A	37.144	37.433	39.318	442.66	36.777	47.834
19	S98T002739_X	-1.1533	.13640	-.44255	431.00	-.61125	10.062
20	S98T002739_AX	4124.4	4008.9	3976.4	4335.9	3843.0	4019.9
21	S98T002712	2.0355	.22111	.02912	481.99	.08953	16.084
22	S98T002712_D	1.2430	.20931	.00273	418.43	-.07154	26.625
23	ICSA	249.40	Q-.00547	-.01177	192.31	-.00165	Q-.00827
24	ICSA	249.10	-.00530	-.01117	192.14	.00539	.01024
25	ICSA	251.23	.44291	-.01231	195.12	.00104	.93524
26	CCV	5.1605	4.9584	5.1096	4.9812	4.9374	5.0324
27	CCB	-.00187	.00002	-.00061	.00587	.00075	.00076

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.1496	5.1057	4.9327	5.1424	4.9136	5.2139
2	ICB	.00529	.00775	-.00212	-.00202	.00343	.00568
3	LLS	.40154	.22032	.19822	.09857	.20684	Q.24418
4	ICSA	.01755	.03265	-.03090	.00391	.04530	-.00026
5	ICSA	.01618	1.0014	-.03408	-.01073	.05240	-.00023
6	PREPBLKTJA	.00923	.01245	-.00633	.00456	-.00142	.01219
7	S98T002844_L	4.4723	.58046	-.76580	-1.4708	1.1529	4.3453
8	S98T002844	5.8758	.72942	2.1563	-.05431	1.3193	3.4616
9	S98T002844_D	6.0836	.21191	2.8118	.02866	1.1311	4.3438
10	S98T002844_A	47.645	40.862	42.415	42.082	40.620	45.830
11	S98T002844_X	7.3402	2.2664	.63678	1.3937	3.8567	3.5834
12	S98T002844_AX	4057.6	4135.7	3967.1	4084.5	3998.3	3986.1
13	CCV	5.0050	5.0731	4.8370	5.0864	4.8127	5.1272
14	CCB	.00777	-.00098	-.00814	-.00966	.01626	.00165
15	S98T002739_L	9.6018	-1.4388	3.2177	-.22422	.79971	4.4822
16	S98T002739	11.162	-.07134	3.0580	.12959	.80038	4.2234
17	S98T002739_D	11.743	.16795	3.3127	.11999	1.2506	4.1235
18	S98T002739_A	49.730	41.081	41.757	39.117	38.262	44.966
19	S98T002739_X	12.874	-2.8603	-.95442	.26112	6.1610	5.7478
20	S98T002739_AX	4010.7	4101.7	3941.5	4041.1	3937.3	3932.5
21	S98T002712	14.472	.61750	21.538	-.07218	1.0533	17.588
22	S98T002712_D	78.390	.23723	14.876	-.08885	.98663	13.352
23	ICSA	.03577	Q-.01313	-.02721	.00999	-.07333	-.00330
24	ICSA	.02881	.03256	-.03194	.00778	.06014	-.00354
25	ICSA	.02956	.99661	-.03632	-.01001	.05582	.00337
26	CCV	5.0977	5.1235	4.9213	5.1549	4.9424	5.1876
27	CCB	.01099	-.00161	-.00604	-.00227	.02449	.00220

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	5.0003	5.0041	.21036	5.1563	4.8416	9.8879
2	ICB	.00292	.00000	-.00373	-.00001	.00612	.00263
3	LLS	.20220	.02022	.00483	.02068	.42816	.47935
4	ICSA	-.00677	.00188	.00653	.00173	.00934	.06461
5	ICSA	-.01093	.00195	.00174	.00104	-.03754	.07319
6	PREPBLKTJA	.01144	.00021	-.00111	.00069	.00141	-.04500

Analysis Report

Averages

Mon 12-21-98 12:43:06 PM

page 5

#	Sample Name	Sm	Sr	Th	Ti	Ti	U
7	S98T002844_L	1.5468	2.1561	-.04432	-.00103	.33413	6.1310
8	S98T002844	.61349	2.0878	-.04546	.04387	-.23680	3.5471
9	S98T002844_D	.92489	2.0370	-.21069	.07923	-.22529	4.4034
10	S98T002844_A	39.777	42.026	1.7794	41.408	39.339	79.924
11	S98T002844_X	.43679	2.1564	-.50818	.26776	.69301	1.4059
12	S98T002844_AX	4023.9	4089.9	175.45	4049.4	4054.4	7787.8
13	CCV	4.8803	4.8971	.21393	5.0560	4.7894	9.6087
14	CCB	.00302	.00000	-.01086	-.00091	.01028	.00800
15	S98T002739_L	1.1486	.06152	-.92534	-.04579	-2.3011	-1.9401
16	S98T002739	.20323	.06856	.11232	.06497	-.42031	-2.5387
17	S98T002739_D	.51313	.07199	.04394	.08252	-.84190	-1.3183
18	S98T002739_A	37.410	36.386	-.35345	37.704	35.778	73.503
19	S98T002739_X	3.1467	.08727	-2.0408	-.09573	1.6798	4.7865
20	S98T002739_AX	3847.7	3919.6	174.22	3919.2	3986.8	7450.7
21	S98T002712	.57724	1.7853	.10324	.28558	-.07551	6.6733
22	S98T002712_D	.24609	1.8750	-.00819	.15983	-.23405	4.6136
23	ICSAB	.00903	.00192	.00580	.00041	.00064	.06234
24	ICSA	-.00658	.00188	.00385	.00151	.01748	.05336
25	ICSAB	-.01281	.00194	.00734	.00085	-.03133	.06108
26	CCV	4.9408	4.9516	.21401	5.1140	4.8725	9.7319
27	CCB	.00658	.00004	-.00418	-.00023	-.00338	.01027

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0516	.00810	4.7690	4.9293
2	ICB	.00084	-.00007	.00021	.00057
3	LLS	.09983	-.00004	.02019	.01961
4	ICSA	.00235	.00812	-.00397	-.00291
5	ICSAB	.46826	.00780	.92949	-.00295
6	PREPBLKTJA	.00243	.00041	.00094	.00161
7	S98T002844_L	.26438	.04559	.00817	.24391
8	S98T002844	.13096	.03705	.02555	.09848
9	S98T002844_D	.21968	.05276	.01451	.18409
10	S98T002844_A	40.517	.05601	38.804	39.338
11	S98T002844_X	-.00282	-.03273	.01326	.03430
12	S98T002844_AX	4019.2	6.4921	3801.3	4024.3
13	CCV	4.9660	.00806	4.7302	4.8370
14	CCB	.00069	-.00008	.00003	.00110
15	S98T002739_L	.10931	-.03131	.24556	.06987
16	S98T002739	.05231	-.00213	.31263	.01826
17	S98T002739_D	.12616	.01984	-.00937	.08387
18	S98T002739_A	38.387	.13587	37.696	36.856
19	S98T002739_X	.68784	.06223	.17574	.64153
20	S98T002739_AX	3933.3	6.6996	3791.7	3915.0
21	S98T002712	.13610	.02608	.04333	.09638
22	S98T002712_D	.16141	.00768	.01168	.06991
23	ICSAB	.00181	.00765	.00417	.00303
24	ICSA	.00225	.00773	-.00445	-.00265
25	ICSAB	.46657	.00757	.92758	-.00364
26	CCV	5.0253	.00817	4.7836	4.8930
27	CCB	.00049	.00007	.00023	.00120

LABCORE Data Entry Template for Worklist# 27099

Analyst: JK 540 Instrument: ICP04 2 11-11-98 Book# 77B30Method: LA-505-151/161 Rev/Mod C-3Worklist Comment: ICP BY-105 (FUSION)Ensure dose rate of 30 cm is ≤ 50 mrem/hr
prior to performing this analysis.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SERDIL	S98T002920 0 F	@ICP-F01	SOLID		
8	SAMPLE	S98T002920 0 F	@ICP-F01	SOLID	98000490	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01						
9	DUP	S98T002920 0 F	@ICP-F01	SOLID		
10	SPK (1ppm)	S98T002920 0 F	@ICP-F01	SOLID		
11	CCV (16ppm) <u>11-11-98</u>		@ICP-QC	QC		
12	CCB		@ICP-QC	QC		
13	SAMPLE	S98T002921 0 F	@ICP-F01	SOLID	98000490	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01						

Data Entry Comments:

uploaded 12-28-98John Wainell

Validated by:

Sam M. Lang12/28/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

A-0004-1

S Type	Sample#	R	A	Test	Matrix	Group#	Project
				SR-F-01 ,	TI-F-01 ,	TL-F-01 ,	U-F-01 , V-F-01 , ZN-F-01 ,
				ZR-F-01			
14 DUP	S98T002921	0	F	@ICP-F01	SOLID		
15 SERDIL	S98T002922	0	F	@ICP-F01	SOLID		
16 SAMPLE	S98T002922	0	F	@ICP-F01	SOLID	98000490	BY-105 (2)
	Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
17 DUP	S98T002922	0	F	@ICP-F01	SOLID		
18 SPK SRAB (10ppm)	S98T002922	0	F	@ICP-F01	SOLID		
19 ICSA SRAB (10ppm) 2/11/98				@ICP-QC	QC		
20 ICSAB				@ICP-QC	QC		
21 CCV				@ICP-QC	QC		
22 CCB				@ICP-QC	QC		

Final page for worklist # 27099

Analyst Signature	Date
JK [Signature]	11-18-98
Peggy K TJG	.25-10 JK 1
SF870027926-L	.25-10-28 JK 205
SF870027920	.25-10 41
SF870027920-D	.25-10 41
SF870027920-E	.25-10 41
SF870027920-X	.25-10-19 410
SF870027920-SX	.25-10-19 410

Analyst Signature	Date
5887061241	.25-10
5887061241	.25-10
5887061241	.25-10-18
5887061241	.25-10
5887061241	.25-10
5887061241	.25-10
5887061241	.25-10-19
5887061241	.25-10-19

Data Entry Comments:

S = Worklist Slot Number, *R* = Replicate Number, *A* = Aliquot Code.

S98T00 2920

$$Al = 3.7338 \frac{\mu g}{ml} \times \frac{1}{2.0416 \frac{g}{L} \times \frac{1L}{1000ml}} = 1.832 + 3 \frac{\mu g}{g}$$

$$\text{Post Spike: } Al = \frac{\left(\frac{42.604}{41} \right) - \left(\frac{3.7338}{41} \right)}{1} \times 100 = 94.8\%$$

$$Na = \frac{\left(\frac{4414.4}{410} \right) - \left(\frac{661.84}{410} \right)}{10} \times 100 = 91.5\%$$

Analysis Report

Summary

Fri 12-18-98 02:23:03 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981218B	ICP2	12/18/98	12:21	DKS	Q	CONC
2	ICB	981218B	ICP2	12/18/98	12:24	DKS	Q	CONC
3	LLS	981218B	ICP2	12/18/98	12:28	DKS	Q	CONC
4	ICSA	981218B	ICP2	12/18/98	12:30	DKS	Q	CONC
5	ICSAB	981218B	ICP2	12/18/98	12:33	DKS	Q	CONC
6	PREPBLKTJA	981218B	ICP2	12/18/98	12:37	DKS	Q	CONC
7	S98T002920_L	981218B	ICP2	12/18/98	12:40	DKS	S	CONC
8	S98T002920_	981218B	ICP2	12/18/98	12:43	DKS	S	CONC
9	S98T002920_D	981218B	ICP2	12/18/98	12:47	DKS	S	CONC
10	S98T002920_A	981218B	ICP2	12/18/98	12:51	DKS	S	CONC
11	S98T002920_X	981218B	ICP2	12/18/98	12:55	DKS	S	CONC
12	S98T002920_AX	981218B	ICP2	12/18/98	12:59	DKS	S	CONC
13	CCV	981218B	ICP2	12/18/98	13:17	DKS	Q	CONC
14	CCB	981218B	ICP2	12/18/98	13:21	DKS	Q	CONC
15	S98T002921	981218B	ICP2	12/18/98	13:24	DKS	S	CONC
16	S98T002921_D	981218B	ICP2	12/18/98	13:27	DKS	S	CONC
17	S98T002922_L	981218B	ICP2	12/18/98	13:30	DKS	S	CONC
18	S98T002922_	981218B	ICP2	12/18/98	13:33	DKS	S	CONC
19	S98T002922_D	981218B	ICP2	12/18/98	13:37	DKS	S	CONC
20	S98T002922_A	981218B	ICP2	12/18/98	13:40	DKS	S	CONC
21	S98T002922_X	981218B	ICP2	12/18/98	13:46	DKS	S	CONC
22	S98T002922_AX	981218B	ICP2	12/18/98	13:49	DKS	S	CONC
23	ICSA	981218B	ICP2	12/18/98	14:10	DKS	Q	CONC
24	ICSAB	981218B	ICP2	12/18/98	14:13	DKS	Q	CONC
25	CCV	981218B	ICP2	12/18/98	14:16	DKS	Q	CONC
26	CCB	981218B	ICP2	12/18/98	14:19	DKS	Q	CONC

JK *[Signature]*

12-18-98

Work/15/ # 27079

BJ-105

5887002920

5887002921

5887002922

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1920 TO 1924.

Analysis Report

Averages

Fri 12-18-98 02:23:03 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9898	4.8463	5.0500	5.0825	4.7150	5.0860
2	ICB	-.00076	-.00711	-.00825	.00046	-.00009	.00004
3	LLS	.02173	.09966	.20182	.10016	.09541	.01041
4	ICSA	.00192	240.55	.03543	-.00527	.00032	.00000
5	ICSAB	.95123	240.05	.03758	-.01085	.44492	.47114
6	PREPBLKTJA	-.00096	-.00516	-.00264	-.00466	.00004	.00005
7	S98T002920_L	-.14457	2.8876	-.08103	-1.3359	-.01948	.00056
8	S98T002920_	.04831	3.7338	.27467	-.07643	.00501	.00560
9	S98T002920_D	.01216	3.6964	.09168	-.11474	.00637	.00020
10	S98T002920_A	36.922	42.604	41.538	40.447	37.175	40.455
11	S98T002920_X	-.25937	5.604	1.7486	-2.0982	.00593	.00050
12	S98T002920_AX	3172.8	3924.7	4098.5	4022.9	3959.8	4018.8
13	CCV	4.9692	4.565	5.0146	5.0075	4.6476	5.0008
14	CCB	.00446	-.00666	-.00302	-.00278	.00000	-.00008
15	S98T002921	.00977	14.523	.15937	-.21014	.00363	-.00161
16	S98T002921_D	.08306	14.957	-.21348	-.13370	.01194	-.00346
17	S98T002922_L	-.16836	33.537	.09400	-.83930	.31959	.02482
18	S98T002922_	.03844	33.548	.18227	-.19251	.31678	.00692
19	S98T002922_D	.09537	29.210	.14923	-.14328	.21035	.00087
20	S98T002922_A	35.880	70.564	39.952	38.746	35.527	38.758
21	S98T002922_X	.18107	33.408	.60349	-3.2203	.37018	.05141
22	S98T002922_AX	3001.5	3999.0	4143.4	4049.7	4038.5	4072.0
23	ICSA	.00859	240.52	.07077	-.00924	.00029	.00000
24	ICSAB	.96155	240.57	.03954	-.01344	.44904	.47167
25	CCV	5.0275	4.8242	5.0787	5.0750	4.7512	5.0835
26	CCB	.00444	-.01022	-.00327	-.00418	.00008	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0005	5.2530	5.1464	4.9275	4.9974	5.0451
2	ICB	-.00594	-.00492	-.00163	-.01609	.00098	.00011
3	LLS	.19745	.24195	.00995	.19688	.04257	.02189
4	ICSA	-.00534	257.32	.00339	-.00173	-.00159	-.00143
5	ICSAB	-.02405	257.95	.94387	.00472	.46193	.46877
6	PREPBLKTJA	-.01129	-.00024	-.00117	-.01096	.00113	.00086
7	S98T002920_L	-2.9296	.84391	-.27933	-2.9076	-.12683	.16039
8	S98T002920_	.32930	.65896	-.00519	.09322	.00719	.37507
9	S98T002920_D	-.17798	.32632	-.02627	-.26270	.07112	.45995
10	S98T002920_A	39.939	42.782	41.631	39.106	40.266	41.062
11	S98T002920_X	-.61055	-.68037	-.19996	-3.6752	.40186	.76714
12	S98T002920_AX	4101.3	4201.4	4082.5	4002.3	4057.2	4113.6
13	CCV	5.0017	5.2759	5.1234	4.8424	4.9752	5.0217
14	CCB	.00715	-.00096	-.00029	-.00539	.00014	.00001
15	S98T002921	.61102	.46486	-.02644	-.27622	.04422	.76525
16	S98T002921_D	-.60030	.41888	.02731	.30506	.05878	.77658
17	S98T002922_L	-.28392	58.129	-.07401	-.86893	.20460	1.4050
18	S98T002922_	1.9514	57.881	.00635	-.05601	.02917	1.5049
19	S98T002922_D	1.5689	52.702	.04344	.41920	.02026	1.5454
20	S98T002922_A	40.229	100.94	41.001	37.147	39.809	41.530
21	S98T002922_X	-.77295	59.802	.21265	1.6851	-.03373	1.5910
22	S98T002922_AX	4163.7	4275.8	4126.9	4067.8	4087.3	4146.4
23	ICSA	-.02415	260.75	.00435	-.00054	.00083	-.00171

Analysis Report

Averages

Fri 12-18-98 02:23:03 PM

page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
24	ICSAB	-.03312	260.71	.95029	.00009	.46351	.47040
25	CCV	5.0253	5.3247	5.1825	4.9228	5.0339	5.0857
26	CCB	-.01669	-.00084	-.00105	-.00705	-.00034	.00005

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.6176	.00896	4.8331	4.9996	4.9533	4.9117
2	ICB	-.00050	.00043	-.00071	-.14384	-.00171	.00071
3	LLS	.01654	.00129	.09829	.49571	.09973	.02162
4	ICSA	.00241	.01396	92.482	.08011	-.00496	.00318
5	ICSAB	.44883	.01532	92.331	-.11873	-.00406	.96603
6	PREPBLKTJA	.00015	.00098	.00342	Q272.66	-.00124	-.00009
7	S98T002920_L	-.18981	.09933	-.00827	11265.	-.30043	-.02108
8	S98T002920	-.01355	.06835	.20333	10708.	.00162	.04631
9	S98T002920_D	.02983	.06521	.21954	10771.	-.04390	.03790
10	S98T002920_A	37.117	.10109	39.007	11014.	39.491	38.362
11	S98T002920_X	-.23486	.46916	-.17367	11105.	-.53189	.37776
12	S98T002920_AX	3697.3	6.9692	3897.9	15044.	3987.0	3927.7
13	CCV	4.5053	.01001	4.7848	4.6899	4.8494	4.6938
14	CCB	-.00102	.00169	-.00105	-.07854	-.00031	.00122
15	S98T002921	-.00021	.06509	.20507	10645.	-.03001	.04210
16	S98T002921_D	-.02463	.07182	.37601	11930.	.04538	.05050
17	S98T002922_L	-.36121	.49233	26.102	12226.	-.18079	.35446
18	S98T002922	-.10965	.17709	25.379	11852.	-.00860	.09435
19	S98T002922_D	-.07215	.21628	14.565	10725.	.07866	.12079
20	S98T002922_A	34.773	.26777	63.375	11712.	37.193	35.318
21	S98T002922_X	-.43408	1.2867	25.659	12060.	.18082	.92080
22	S98T002922_AX	3756.9	6.7358	3966.4	16459.	4052.5	3947.5
23	ICSA	.00254	.00857	93.071	-.00796	-.00489	.00338
24	ICSAB	.44766	.01178	92.905	.08043	-.00435	.95371
25	CCV	4.6004	.00969	4.8373	4.7782	4.9510	4.7826
26	CCB	-.00049	.00103	-.00020	-.11321	-.00092	.00051

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1169	4.9382	5.0599	4.8823	4.8008	4.9865
2	ICB	-.01202	.00028	.00306	-.01305	-.00145	-.00046
3	LLS	.20513	.02073	.10293	.18836	.19706	.04155
4	ICSA	247.08	-.00457	-.01448	188.77	-.00335	-.00802
5	ICSAB	245.78	.44261	-.01439	187.77	.00021	.91012
6	PREPBLKTJA	-.00324	.00183	.00028	Q.13750	.00310	Q.19359
7	S98T002920_L	-.15246	.06674	.33037	520.75	-.36033	7.5287
8	S98T002920	.05184	.06597	-.03335	495.30	-.04508	8.0258
9	S98T002920_D	-.03734	.06358	.03217	502.70	.10283	10.046
10	S98T002920_A	40.609	39.231	41.013	548.78	38.374	47.968
11	S98T002920_X	-3.1357	.18889	-.12452	661.84	.59386	9.4481
12	S98T002920_AX	4127.4	4075.0	4041.2	4414.4	3866.1	4072.5
13	CCV	4.9999	4.9163	5.0176	4.6790	4.6926	4.9665
14	CCB	.00155	.00062	.00373	-.02119	-.00056	.00031
15	S98T002921	-.02373	.07304	.07194	469.65	.01202	14.854
16	S98T002921_D	.32366	.08631	.05897	466.07	-.05447	16.196
17	S98T002922_L	.71401	.52398	-.42769	357.92	-.03562	23.720

Analysis Report		Averages		Fri 12-18-98 02:23:03 PM			page 4
#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
18	S98T002922	1.4598	.41098	.03875	348.97	.17483	23.216
19	S98T002922_D	1.4681	.31415	-.00693	380.30	.02473	24.619
20	S98T002922_A	40.461	38.885	40.062	380.94	35.969	63.286
21	S98T002922_X	4.2826	.77996	.37729	333.94	.50498	23.874
22	S98T002922_AX	4185.4	4109.4	4083.1	4312.4	3932.9	4142.6
23	ICSA	246.78	-.00439	-.01159	186.02	-.00262	-.00287
24	ICSAB	246.36	.44514	-.01454	186.42	-.00284	.92127
25	CCV	5.0684	4.9711	5.0912	4.7643	4.7904	5.0294
26	CCB	-.00009	.00034	.00532	-.00457	-.00128	-.00115

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0479	5.0845	4.9407	5.0766	4.8481	5.1978
2	ICB	.00120	-.00276	.01604	.00493	.00878	.00996
3	LLS	.40641	.20805	.20828	.10123	.20449	Q.25353
4	ICSA	.02815	.02005	-.02828	.00909	.06617	.00335
5	ICSAB	.05486	1.0042	-.03855	-.00429	.06802	.00165
6	PREPBLKTJA	.01017	.00756	.00867	-.00147	.01482	.00969
7	S98T002920_L	-.72069	-.01839	2.7537	.29856	1.4329	1.8116
8	S98T002920	.73204	-.54448	2.4317	.15201	1.0523	1.0252
9	S98T002920_D	.98630	.26956	2.6357	.24574	.34451	1.0535
10	S98T002920_A	41.745	40.893	43.079	41.294	39.612	43.452
11	S98T002920_X	4.3686	-2.2045	4.5990	-.15858	6.4499	4.5472
12	S98T002920_AX	4016.9	4132.6	3997.1	4061.4	3945.5	4015.2
13	CCV	4.9628	5.0414	4.9214	5.0288	4.7785	5.1614
14	CCB	.00553	-.00135	.00663	.00551	.01814	.00693
15	S98T002921	2.1787	-.01413	1.9380	.15859	.48230	1.3271
16	S98T002921_D	1.4933	-.24166	2.5506	.16537	.52198	1.4813
17	S98T002922_L	21.486	-1.3798	7.6581	.56937	2.6924	12.361
18	S98T002922	21.887	.61474	5.5302	-.00241	1.2596	10.455
19	S98T002922_D	16.475	-.00713	7.0514	.47050	1.1996	11.967
20	S98T002922_A	62.195	41.357	45.575	40.720	39.982	52.679
21	S98T002922_X	24.690	-3.2000	10.258	2.0909	6.9350	15.348
22	S98T002922_AX	4096.3	4160.8	4046.2	4103.0	4014.7	4051.6
23	ICSA	.04297	.02535	-.02979	.00781	.08271	-.00174
24	ICSAB	.04694	1.0048	-.03000	.00955	.06890	.00476
25	CCV	5.0944	5.1245	4.9671	5.1136	4.8817	5.2212
26	CCB	.00690	-.00594	-.00022	.00546	.02036	.00296

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.7956	4.8503	.23800	5.0360	4.8045	9.4707
2	ICB	.00537	.00008	-.00307	-.00001	.00038	.02181
3	LLS	.20358	.01986	.01137	.02029	.40567	.50814
4	ICSA	-.00548	.00167	.00602	.00092	.00090	-.04915
5	ICSAB	-.01327	.00169	.01485	.00184	.01436	-.06885
6	PREPBLKTJA	.00728	.00017	.00277	.00026	-.00199	-.05132
7	S98T002920_L	.60352	.02594	-.43173	-.13784	1.5071	.93653
8	S98T002920	.63655	.02609	.25831	-.01594	-.51358	-.69831
9	S98T002920_D	.63920	.02958	.11404	.02907	.74042	-.75542
10	S98T002920_A	38.306	38.572	1.8288	40.461	38.612	73.745
11	S98T002920_X	4.0267	.05076	-.88804	-.00547	-5.6976	15.639

Analysis Report

Averages

Fri 12-18-98 02:23:03 PM

page 5

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
12	S98T002920_AX	3846.6	3953.0	183.87	3990.3	3995.8	7449.6
13	CCV	4.6724	4.7401	.22105	4.9694	4.7201	9.2224
14	CCB	.01640	.00029	-.00159	-.00023	.00793	.06629
15	S98T002921	.62460	.02262	.00525	.02857	.35590	-.62427
16	S98T002921_D	.73463	.02434	.30634	.02069	.33962	-.46900
17	S98T002922_L	4.8656	11.972	-.34925	.11290	3.4340	57.653
18	S98T002922	1.1127	11.638	.92362	.14632	.04343	44.408
19	S98T002922_D	1.7139	6.7840	.59398	.12993	.86006	22.347
20	S98T002922_A	36.951	47.826	2.4496	38.901	37.514	114.89
21	S98T002922_X	11.537	11.978	-.72583	.42506	6.6740	86.831
22	S98T002922_AX	3905.7	4025.0	188.34	4042.6	4032.0	7613.2
23	ICSA	-.00442	.00168	.00503	.00137	-.00509	-.06008
24	ICSAB	-.00774	.00168	.00267	.00114	.00840	-.06377
25	CCV	4.7747	4.8325	.22461	5.0506	4.8221	9.3987
26	CCB	.00948	.00008	-.00082	-.00022	.00869	.03324

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9815	.00851	4.8064	4.8148
2	ICB	.00082	.00030	-.00043	.00111
3	LLS	.10207	.00095	.01939	.02014
4	ICSA	.00321	.00824	-.00380	-.00263
5	ICSAB	.46275	.00810	.93517	-.00369
6	PREPBLKTJA	.00146	.00033	-.00042	.00104
7	S98T002920_L	.21130	.06386	-.06495	.15582
8	S98T002920	.13489	.03564	.01137	.06805
9	S98T002920_D	.12113	.03566	-.00361	.09113
10	S98T002920_A	40.147	.08433	39.359	38.536
11	S98T002920_X	1.0350	.28254	-.06467	.79672
12	S98T002920_AX	3994.2	6.8270	3874.0	3954.4
13	CCV	4.9316	.00916	4.8162	4.7478
14	CCB	.00367	.00091	.00015	.00278
15	S98T002921	.13524	.03560	-.02274	.09918
16	S98T002921_D	.16242	.04196	.00272	.07660
17	S98T002922_L	.91615	.28075	.21307	.65675
18	S98T002922	.27181	.08394	.10761	.14008
19	S98T002922_D	.36403	.10617	.06269	.20792
20	S98T002922_A	39.327	.17288	39.650	37.122
21	S98T002922_X	2.5947	.70213	.10215	1.7115
22	S98T002922_AX	4034.6	6.7896	3890.8	4011.4
23	ICSA	.00374	.00841	-.00396	-.00172
24	ICSAB	.46618	.00833	.94283	-.00195
25	CCV	4.9978	.00885	4.8534	4.8302
26	CCB	.00148	.00053	-.00026	.00154

JK

12-18-98

LABCORE Data Entry Template for Worklist# 27100

Analyst: JK 5670 Instrument: ICP01 2 11-11-98 Book# 77830Method: LA-505-151/161 Rev/Mod C-3Worklist Comment: ICP BY-105 (FUSION)Ensure dose rate at 30 cm is ≤ 50 mrem/hr
prior to performing this analysis.

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SERDIL	S98T002986 0 F	@ICP-F01	SOLID		
8	SAMPLE	S98T002986 0 F	@ICP-F01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01, MN-F-01, MO-F-01, NA-F-01, ND-F-01, NI-F-01, P-F-01, PB-F-01, S-F-01, SB-F-01, SE-F-01, SI-F-01, SM-F-01, SR-F-01, TI-F-01, TL-F-01, U-F-01, V-F-01, ZN-F-01, ZR-F-01						
9	DUP	S98T002986 0 F	@ICP-F01	SOLID		
10	SPK (10ppm)	S98T002986 0 F	@ICP-F01	SOLID		
11	CCV		@ICP-QC	QC		
12	CCB		@ICP-QC	QC		
13	SERDIL	S98T002987 0 F	@ICP-F01	SOLID		
14	SAMPLE	S98T002987 0 F	@ICP-F01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-F-01, AL-F-01, AS-F-01, B-F-01, BA-F-01, BE-F-01, BI-F-01, CA-F-01, CD-F-01, CE-F-01, CO-F-01, CR-F-01, CU-F-01, FE-F-01, LA-F-01, LI-F-01, MG-F-01,						

Data Entry Comments:

Uploaded & Validated by:
Saul H. Pang
12/15/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

11/10/98 13:05
A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27100

S Type	Sample#	R A	Test	Matrix	Group#	Project
			MN-F-01 ,	MO-F-01 ,	NA-F-01 ,	ND-F-01 , NI-F-01 , P-F-01 ,
			PB-F-01 ,	S-F-01 ,	SB-F-01 ,	SE-F-01 , SI-F-01 , SM-F-01 ,
			SR-F-01 ,	TI-F-01 ,	TL-F-01 ,	U-F-01 , V-F-01 , ZN-F-01 ,
			ZR-F-01			
15 DUP	S98T002987	0 F	@ICP-F01	SOLID		
16 SPK, (1/10/98)	S98T002987	0 F	@ICP-F01	SOLID		
17 ICSA	SPK (10/10/98) 11-11-98		@ICP-QC	QC		
18 ICSAB			@ICP-QC	QC		
19 CCV			@ICP-QC	QC		
20 CCB			@ICP-QC	QC		

Final page for worklist # 27100

DK LDO 11-15-98

Analyst Signature	Date
Procs HJG	11-10-98
5987001986-L	11-10-98
5987001986	11-10-98
5987001986-D	11-10-98
5987001986-G	11-10-98
5987001986-Y	11-10-98
5987001986-GK	11-10-98
5987001987-L	11-10-98
5987001987	11-10-98
5987001987-D	11-10-98
5987001987-G	11-10-98
5987001987-X	11-10-98
5987001987-GK	11-10-98

Analyst Signature	Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

Summary

Tue 12-15-98 10:55:20 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981215A	ICP2	12/15/98	09:11	DKS	Q	CONC
2	ICB	981215A	ICP2	12/15/98	09:14	DKS	Q	CONC
3	LLS	981215A	ICP2	12/15/98	09:17	DKS	Q	CONC
4	ICSA	981215A	ICP2	12/15/98	09:20	DKS	Q	CONC
5	ICSAB	981215A	ICP2	12/15/98	09:24	DKS	Q	CONC
6	PREPBLKTJA	981215A	ICP2	12/15/98	09:28	DKS	Q	CONC
7	S98T002986_L	981215A	ICP2	12/15/98	09:31	DKS	S	CONC
8	S98T002986	981215A	ICP2	12/15/98	09:34	DKS	S	CONC
9	S98T002986_D	981215A	ICP2	12/15/98	09:37	DKS	S	CONC
10	S98T002986_A	981215A	ICP2	12/15/98	09:41	DKS	S	CONC
11	S98T002986_X	981215A	ICP2	12/15/98	09:46	DKS	S	CONC
12	S98T002986_AX	981215A	ICP2	12/15/98	09:49	DKS	S	CONC
13	CCV	981215A	ICP2	12/15/98	10:01	DKS	Q	CONC
14	CCB	981215A	ICP2	12/15/98	10:04	DKS	Q	CONC
15	S98T002987_L	981215A	ICP2	12/15/98	10:07	DKS	S	CONC
16	S98T002987	981215A	ICP2	12/15/98	10:10	DKS	S	CONC
17	S98T002987_D	981215A	ICP2	12/15/98	10:13	DKS	S	CONC
18	S98T002987_A	981215A	ICP2	12/15/98	10:17	DKS	S	CONC
19	S98T002987_X	981215A	ICP2	12/15/98	10:22	DKS	S	CONC
20	S98T002987_AX	981215A	ICP2	12/15/98	10:26	DKS	S	CONC
21	ICSA	981215A	ICP2	12/15/98	10:42	DKS	Q	CONC
22	ICSAB	981215A	ICP2	12/15/98	10:45	DKS	Q	CONC
23	CCV	981215A	ICP2	12/15/98	10:48	DKS	Q	CONC
24	CCB	981215A	ICP2	12/15/98	10:52	DKS	Q	CONC

JK

12-15-98

Work list # 27100

84-105

598T002986

598T002987

598T002986

$$al = 71.372 \frac{\mu g}{ml} \times \frac{1}{1.972 \frac{g}{L} \times \frac{1L}{100ml}} = 3.622 + 4 \frac{\mu g}{g}$$

Post Spike:

$$al = \left(\frac{109.82}{41} \right) - \left(\frac{71.372}{41} \right) \times 100 = 93.8\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1927 TO 1931.

Analysis Report

Averages

Tue 12-15-98 10:55:20 AM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0153	4.9164	5.0707	5.1631	4.8701	5.0850
2	ICB	.00065	-.00325	.00737	.00328	.00011	.00008
3	LLS	.02268	.10573	.21785	.10038	.09856	.01041
4	ICSA	.00169	245.36	.07919	.00045	.00032	.00005
5	ICSAB	.95117	243.07	.09229	-.00576	.46042	.47615
6	PREPBLKTJA	.00156	.01068	.01560	-.00281	.00014	.00022
7	S98T002986_L	.47394	74.154	4.2346	-.47651	.13472	.03557
8	S98T002986	.04027	71.372	.16417	-.09371	.09531	.00732
9	S98T002986_D	.10777	88.755	.56232	-.09414	.08182	.00188
10	S98T002986_A	37.807	109.82	41.838	40.705	37.640	40.562
11	S98T002986_X	.50822	75.007	3.3322	-.76076	.13555	.07146
12	S98T002986_AX	3816.2	4048.4	4130.4	4084.7	4045.1	4070.7
13	CCV	4.9861	4.8515	5.0447	5.1274	4.7962	5.0883
14	CCB	.00329	.00447	.01426	-.00045	.00023	.00017
15	S98T002987_L	.14269	75.430	2.8497	-.56767	.06094	.03565
16	S98T002987	.06280	73.755	.12580	-.05110	.04654	.00727
17	S98T002987_D	.07348	89.235	.75773	.07963	.04265	.00732
18	S98T002987_A	37.652	114.35	42.734	41.055	37.894	40.767
19	S98T002987_X	.06125	7.8392	.25571	-.19108	.00947	.00892
20	S98T002987_AX	360.79	400.05	409.47	402.93	398.06	400.20
21	ICSA	.00490	235.84	.04030	-.01057	.00041	.00026
22	ICSAB	.94318	236.47	-.00107	-.00640	.44194	.46010
23	CCV	4.9713	4.7971	5.0102	5.0580	4.7265	4.9973
24	CCB	.00227	.00092	.00425	-.00046	.00018	.00004

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0579	5.2393	5.1352	5.0266	5.0140	5.0332
2	ICB	.01134	.01480	.00085	.00975	.00025	.00175
3	LLS	.22028	.23981	.01122	.21248	.04013	.02144
4	ICSA	.03915	286.14	.00521	.02452	-.00410	-.00477
5	ICSAB	.01772	256.45	.94648	.01406	.46314	.46713
6	PREPBLKTJA	.01375	.00438	.00103	.01193	-.00053	.00029
7	S98T002986_L	9.2882	7.7977	.30958	4.1862	.02099	1.2298
8	S98T002986	1.1210	7.2872	.01908	.26656	-.03704	.95788
9	S98T002986_D	.45885	4.7789	.10384	.82282	-.06331	.96630
10	S98T002986_A	40.321	50.128	42.057	39.869	40.781	41.891
11	S98T002986_X	5.9176	7.8384	.35897	4.4159	.00534	1.1978
12	S98T002986_AX	4129.7	4276.0	4142.0	4066.4	4113.5	4168.1
13	CCV	5.0444	5.2601	5.1540	4.9598	5.0236	5.0458
14	CCB	.04081	.01423	.00027	.01143	.00003	.00191
15	S98T002987_L	4.1161	4.7163	.11761	2.7342	-.07247	1.3595
16	S98T002987	1.3253	4.7915	.03644	.55980	-.02885	1.3294
17	S98T002987_D	1.6571	3.0682	.03589	.19091	.10029	1.3815
18	S98T002987_A	40.960	47.668	42.220	40.228	41.019	42.468
19	S98T002987_X	.57605	.51728	.03102	.42837	.02541	.15745
20	S98T002987_AX	409.40	424.20	409.71	399.16	408.31	412.69
21	ICSA	.00949	256.59	.00575	.02157	-.00045	-.00149
22	ICSAB	-.00010	255.42	.93113	.01569	.45574	.46207
23	CCV	4.9797	5.2520	5.0980	4.8834	4.9756	5.0099
24	CCB	.02845	.01468	.00107	.00781	-.00058	.00065

Analysis Report

Averages

Tue 12-15-98 10:55:20 AM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7422	.00886	4.8506	5.0422	5.0439	4.9891
2	ICB	.00001	.00126	.00098	.25176	.00117	.00062
3	LLS	.01781	.00076	.09843	.62443	.10318	.02142
4	ICSA	.00126	-.01509	101.07	.23074	-.00358	.00379
5	ICSAB	.46151	.02372	92.946	.24884	-.00278	.99025
6	PREPBLKTJA	.00015	.00141	.00442	Q269.30	.00215	.00084
7	S98T002986_L	.07956	.37253	6.8687	12175.	.60574	.21515
8	S98T002986_	.02907	.07422	6.6557	12030.	.01850	.05610
9	S98T002986_D	.03542	.10073	4.9930	11221.	.08622	.06468
10	S98T002986_A	37.374	.17529	45.596	11810.	39.566	37.983
11	S98T002986_X	-.13377	.98315	6.1405	12128.	.31643	.73135
12	S98T002986_AX	3760.5	7.0493	3956.7	16062.	4034.6	3965.9
13	CCV	4.6673	.00959	4.8280	4.9470	4.9640	4.8511
14	CCB	.00001	.00227	-.00018	.29946	.00143	.00178
15	S98T002987_L	-.05582	.46236	16.792	10799.	.23789	.34420
16	S98T002987_	.04330	.10939	16.944	10562.	.05434	.08618
17	S98T002987_D	.04584	.11616	7.2548	10522.	.04334	.07755
18	S98T002987_A	37.573	.16363	56.161	10757.	39.789	38.384
19	S98T002987_X	.00165	.08853	1.6765	1089.0	.05238	.06883
20	S98T002987_AX	370.52	.73319	392.41	1460.6	396.26	389.05
21	ICSA	.00197	.01105	91.421	.25857	-.00232	.00451
22	ICSAB	.43990	.01636	91.087	.26262	-.00266	.93956
23	CCV	4.5954	.01050	4.7645	4.8928	4.8872	4.7704
24	CCB	-.00032	.00123	.00019	.07211	.00139	.00104

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.2190	4.9814	5.1008	4.9664	4.9234	5.0033
2	ICB	.00747	.00042	.00745	.01284	.00505	-.00369
3	LLS	.21517	.02049	.10560	.20520	.20265	.03914
4	ICSA	249.35	-.00341	-.01621	187.41	-.00747	-.00522
5	ICSAB	249.67	.44394	-.01492	192.09	-.00053	.91086
6	PREPBLKTJA	.01749	.00170	.00108	.08206	.00101	Q.13870
7	S98T002986_L	6.1963	.33012	.29486	384.65	.88341	16.013
8	S98T002986_	2.0400	.23719	.03169	379.13	.09981	16.075
9	S98T002986_D	2.1080	.23524	.03850	382.37	.13761	9.9022
10	S98T002986_A	43.186	39.622	41.430	410.36	38.558	56.161
11	S98T002986_X	5.5238	.54825	.14448	372.58	1.3096	16.251
12	S98T002986_AX	4185.7	4140.2	4101.6	4320.8	3927.7	4149.3
13	CCV	5.1131	4.9548	5.0864	4.8013	4.8251	5.0142
14	CCB	.01479	.00087	.00415	-.00572	.00424	-.00439
15	S98T002987_L	1.9238	.41206	.33193	405.28	.17384	10.009
16	S98T002987_	1.3193	.29518	.06507	400.18	.16270	10.291
17	S98T002987_D	.93395	.22161	.05840	405.05	.29304	8.8437
18	S98T002987_A	42.549	39.861	41.749	444.87	38.575	51.738
19	S98T002987_X	.61485	.05729	.03394	40.345	.08693	.92493
20	S98T002987_AX	414.47	409.29	406.65	428.17	385.08	409.48
21	ICSA	243.21	-.00417	-.01501	182.22	.00084	-.00369
22	ICSAB	243.53	.43679	-.01253	183.27	.00161	.90175
23	CCV	5.0429	4.9149	5.0548	4.7198	4.7464	4.9675
24	CCB	.01053	.00042	.00429	.01105	.00217	.00523

Analysis Report

Averages

Tue 12-15-98 10:55:20 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.1359	5.0749	4.9563	5.1336	4.8559	5.2226
2	ICB	.00331	.00751	.01138	-.00181	.00403	.00960
3	LLS	.39251	.21063	.20641	.09770	.19094	Q.24871
4	ICSA	.09475	.01089	-.05367	.01188	.11988	.00963
5	ICSAB	.03031	.98322	-.02921	.00906	.04668	.00016
6	PREPBLKTJA	.00738	-.00627	.01435	-.00064	.01938	.02575
7	S98T002986_L	3.0972	.81650	3.8422	1.1368	3.1965	61.356
8	S98T002986	3.0086	.29125	3.3368	-.05079	.42107	58.143
9	S98T002986_D	3.8272	.05071	3.3727	.12367	.74120	34.440
10	S98T002986_A	44.791	41.977	43.664	41.698	41.096	100.87
11	S98T002986_X	3.3869	.23813	6.4536	3.4942	11.602	67.809
12	S98T002986_AX	4075.3	4202.6	4046.5	4125.2	4005.8	4109.9
13	CCV	5.1247	5.0908	4.9178	5.1391	4.8779	5.2022
14	CCB	.00646	-.00226	.00229	-.00227	.01741	.01446
15	S98T002987_L	11.176	-1.7413	3.9832	-.05525	3.5185	19.080
16	S98T002987	10.890	.20600	3.9492	.11259	1.2568	14.877
17	S98T002987_D	6.3411	.45114	5.1737	-.14750	1.2453	15.420
18	S98T002987_A	53.018	41.862	44.976	42.417	41.154	58.142
19	S98T002987_X	1.2218	-.01894	.70325	-.18408	.84409	2.3892
20	S98T002987_AX	405.96	417.28	403.68	409.10	395.72	404.39
21	ICSA	.04352	.02375	-.03028	.00167	.08087	.00876
22	ICSAB	.05881	1.0141	-.03021	.00230	.07874	.00836
23	CCV	4.9857	5.0870	4.9288	5.0735	4.8443	5.1754
24	CCB	.00076	-.00596	.00687	-.00123	.02939	.00917

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8989	4.9204	.21194	5.1092	4.8197	9.7252
2	ICB	.00976	.00016	.00158	.00067	-.00384	.03639
3	LLS	.20244	.02020	.01106	.02059	.41120	.49565
4	ICSA	-.00804	.00152	.03430	.00174	-.05115	-.12162
5	ICSAB	-.01083	.00191	.00799	.00155	-.01735	.03366
6	PREPBLKTJA	.00963	.00021	.00935	-.00015	.00441	-.03363
7	S98T002986_L	3.2942	.21618	.70324	1.0538	.15826	8.6158
8	S98T002986	.66912	.17184	.24674	.90706	-.38669	-.76305
9	S98T002986_D	.79417	.16665	.38266	.65147	.47286	.00698
10	S98T002986_A	38.853	38.713	2.0554	41.471	38.903	75.808
11	S98T002986_X	8.4899	.29229	1.7256	1.1905	12.284	32.264
12	S98T002986_AX	3893.0	4004.5	172.84	4036.9	4035.5	7568.9
13	CCV	4.8088	4.8474	.21844	5.0668	4.7972	9.5584
14	CCB	.02246	.00037	.00201	.00066	.01251	.08602
15	S98T002987_L	4.2471	.26791	.07882	.82166	.90239	14.393
16	S98T002987	.96827	.20999	.30033	.74180	.63154	1.1402
17	S98T002987_D	1.1178	.21872	.11907	.34870	.32518	1.3022
18	S98T002987_A	39.167	38.932	2.2518	41.673	39.189	77.233
19	S98T002987_X	.86449	.03446	-.04082	.10005	-.09473	3.0987
20	S98T002987_AX	381.81	393.34	17.043	398.33	400.00	743.60
21	ICSA	.01164	.00177	.01170	.00203	-.00073	.05766
22	ICSAB	.00757	.00186	.01115	.00202	-.03619	.06901
23	CCV	4.7334	4.7721	.21455	4.9977	4.7306	9.3841
24	CCB	.01087	.00020	.00043	.00066	.01326	.03732

Analysis Report

Averages

Tue 12-15-98 10:55:20 AM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0299	.00895	4.8047	4.8934
2	ICB	.00248	.00092	-.00068	.00165
3	LLS	.10152	.00080	.01936	.01924
4	ICSA	.00488	.00945	-.00400	-.00279
5	ICSAB	.46710	.00821	.93166	-.00269
6	PREPBLKTJA	.00277	.00087	.00063	.00144
7	S98T002986_L	.82162	.25339	-.15577	.49539
8	S98T002986	.15996	.04852	.01473	.10536
9	S98T002986_D	.22890	.06104	-.01896	.10699
10	S98T002986_A	40.617	.13807	39.809	38.933
11	S98T002986_X	2.1807	.69318	.00145	1.2036
12	S98T002986_AX	4052.8	7.2581	3938.2	4007.9
13	CCV	5.0071	.00933	4.8318	4.8476
14	CCB	.00530	.00168	-.00017	.00339
15	S98T002987_L	1.0920	.31519	-.05954	.66121
16	S98T002987	.25961	.07341	.01031	.16134
17	S98T002987_D	.26911	.07334	-.01213	.16249
18	S98T002987_A	40.794	.14068	39.981	39.041
19	S98T002987_X	.21154	.05973	-.02277	.14880
20	S98T002987_AX	400.74	.73146	392.50	395.52
21	ICSA	.00658	.00945	-.00342	-.00006
22	ICSAB	.46144	.00937	.93380	-.00011
23	CCV	4.9583	.00964	4.8220	4.7846
24	CCB	.00266	.00100	-.00070	.00174

JK

12-15-98

LABCORE Data Entry Template for Worklist# 27101

Analyst:

*JK*Instrument: ICPOT-1 *11-11-98*

Book#

*77830*Method: LA-505-151/161 Rev/Mod *C-3**Ensure data rate at 30 cm is ≤ 50 kcps/hr
prior to performing this analysis.*Worklist Comment: *11-11-98* ICP BY-105 (FUSION)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPBLKTJA		@ICP-F01	SOLID		
7	SAMPLE	S98T002988 0 F	@ICP-F01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
8	DUP	S98T002988 0 F	@ICP-F01	SOLID		
9	SAMPLE	S98T002989 0 F	@ICP-F01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
10	DUP	S98T002989 0 F	@ICP-F01	SOLID		
11	ICSA		@ICP-QC	QC		
12	ICSAB		@ICP-QC	QC		

Data Entry Comments:

*uploaded 12-14-98
John Warrall**Validated by:
Saul N. Pang
12/15/98*

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27101

S Type	Sample#	R A	Test	Matrix	Group#	Project
13	CCV		@ICP-QC	QC		
14	CCB		@ICP-QC	QC		

Final page for worklist # 27101

HA 11-14-98

Analyst Signature	Date
Prepb/KJG	.15-10, DF 1
S987002988	.15-10, DF 41
S987002988-D	.15-10 41
S987002989	.15-10, DF 41
S987002989-D	.15-10 41

Analyst Signature	Date

S987002988

$$AL = 67.780 \frac{\mu g}{mL} \times \frac{1}{1.9872 \frac{g}{L} \times \frac{16}{1000000}} = 3.41e+4 \frac{\mu g}{g}$$

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

Summary

Mon 12-14-98 10:04:54 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981214A	ICP2	12/14/98	09:11	DKS	Q	CONC
2	ICB	981214A	ICP2	12/14/98	09:15	DKS	Q	CONC
3	LLS	981214A	ICP2	12/14/98	09:18	DKS	Q	CONC
4	ICSA	981214A	ICP2	12/14/98	09:21	DKS	Q	CONC
5	ICSAB	981214A	ICP2	12/14/98	09:24	DKS	Q	CONC
6	PREPBLKTJA	981214A	ICP2	12/14/98	09:27	DKS	Q	CONC
7	S98T002988	981214A	ICP2	12/14/98	09:37	DKS	S	CONC
8	S98T002988_D	981214A	ICP2	12/14/98	09:41	DKS	S	CONC
9	S98T002989	981214A	ICP2	12/14/98	09:44	DKS	S	CONC
10	S98T002989_D	981214A	ICP2	12/14/98	09:47	DKS	S	CONC
11	ICSA	981214A	ICP2	12/14/98	09:52	DKS	Q	CONC
12	ICSAB	981214A	ICP2	12/14/98	09:55	DKS	Q	CONC
13	CCV	981214A	ICP2	12/14/98	09:58	DKS	Q	CONC
14	CCB	981214A	ICP2	12/14/98	10:01	DKS	Q	CONC

JK

12-14-98

Work/15/ # 27101

89-05

5987002988

5987002989

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1934 TO 1937.

Analysis Report

Averages

Mon 12-14-98 10:04:54 AM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0094	4.8549	5.0126	5.1628	4.8490	5.0738
2	ICB	-.00138	-.00614	.00397	.00367	.00002	.00012
3	LLS	.01987	.09575	.19751	.10108	.09751	.01025
4	ICSA	.00081	238.72	.06234	-.00592	.00022	.00030
5	ICSAB	.94760	239.32	.02633	-.00295	.45419	.46942
6	PREPBLKTJA	-.00044	.00029	.00253	-.00092	.00007	.00000
7	S98T002988	-.05602	67.780	-.36710	-.18563	.12214	.00728
8	S98T002988_D	-.08599	61.657	.14635	-.18602	.07058	-.00157
9	S98T002989	-.03961	49.868	-.00749	.03867	.05643	.00371
10	S98T002989_D	-.04977	52.890	-.02331	-.07416	.05986	.00375
11	ICSA	.00105	237.01	.02919	-.00604	.00030	.00013
12	ICSAB	.94279	237.75	.04133	-.00637	.44884	.46813
13	CCV	4.9982	4.8324	5.0567	5.1623	4.8506	5.1143
14	CCB	-.00011	-.00093	.00130	.00091	.00013	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0348	5.2235	5.1223	5.0074	5.0156	5.0222
2	ICB	-.00277	.00795	-.00128	-.00129	-.00155	-.00112
3	LLS	.19571	.23015	.00889	.19177	.04003	.01854
4	ICSA	.00549	252.47	.00361	-.00079	-.00086	-.00351
5	ICSAB	-.02471	256.30	.93517	.00831	.45902	.46253
6	PREPBLKTJA	-.00562	.01729	-.00108	-.00223	-.00054	-.00102
7	S98T002988	-.44423	6.7670	-.06414	-.63713	-.14257	.83135
8	S98T002988_D	-.56864	4.5901	-.07273	-.45527	-.12886	.55387
9	S98T002989	-.68271	5.8978	-.07319	-.33610	-.02945	1.0521
10	S98T002989_D	-.14180	4.7388	-.05831	-.54608	-.10716	.97381
11	ICSA	-.01825	253.66	.00423	-.00226	.00153	-.00322
12	ICSAB	-.01819	254.24	.93161	-.00128	.45496	.46005
13	CCV	5.0761	5.1978	5.1285	5.0332	5.0222	5.0308
14	CCB	-.01813	.01035	-.00060	.00411	.00024	-.00125

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7290	.00938	4.8457	4.6399	5.0277	4.9812
2	ICB	-.00078	-.00019	-.00021	-.27133	-.00014	-.00050
3	LLS	.01685	-.00013	.09630	Q.34305	.10025	.02074
4	ICSA	.00254	.03296	91.817	-.12436	-.00389	.00181
5	ICSAB	.45712	.03081	92.244	-.06277	-.00345	.97283
6	PREPBLKTJA	-.00011	.00089	.00210	Q262.00	-.00015	.00020
7	S98T002988	.01883	.01735	7.8129	11465.	-.07415	-.00398
8	S98T002988_D	-.01240	.03024	6.6928	10925.	-.05550	.00426
9	S98T002989	.02939	.05756	3.2100	10635.	-.02789	.02900
10	S98T002989_D	-.00667	.02274	3.8623	10887.	-.04272	.02080
11	ICSA	.00250	.02858	91.458	-.10645	-.00419	.00252
12	ICSAB	.45321	.03198	91.637	-.09488	-.00458	.96022
13	CCV	4.7529	.00983	4.8166	4.6560	5.0516	4.9858
14	CCB	-.00076	.00042	.00031	-.12720	.00072	.00010

Analysis Report

Averages

Mon 12-14-98 10:04:54 AM

page 3

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1803	4.9656	5.0949	4.9520	4.9047	5.0008
2	ICB	-.00310	-.00008	.00236	-.00828	-.00234	-.00580
3	LLS	.20177	.01994	.10195	.19394	.19746	.03524
4	ICSA	247.73	-.00505	-.01392	190.45	-.00351	-.00563
5	ICSAB	247.58	.44126	-.01540	189.35	-.00236	.91179
6	PREPBLKTJA	.00137	.00118	.00123	.04597	-.00075	Q.14843
7	S98T002988	1.3424	.21404	-.00789	422.64	-.11163	8.8758
8	S98T002988_D	.74194	.19864	.00529	426.88	-.15774	11.487
9	S98T002989	.40992	.14479	-.00729	421.29	-.01510	10.788
10	S98T002989_D	.44226	.14526	.01161	433.59	-.11584	7.7038
11	ICSA	245.23	-.00488	-.01424	186.59	-.00191	-.01000
12	ICSAB	245.49	.43801	-.01793	186.85	-.00296	.90727
13	CCV	5.1250	4.9474	5.0959	4.9095	4.9228	4.9899
14	CCB	.00007	.00014	.00363	-.01244	-.00167	-.00167

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0567	5.0481	4.9248	5.0799	4.8667	5.2400
2	ICB	-.00233	-.01578	.00801	-.00261	.00283	.00547
3	LLS	.38663	.19710	.19521	.09282	.19444	Q.24900
4	ICSA	.02212	.02091	-.01536	.00435	.05844	.00545
5	ICSAB	.02179	.98975	-.01769	.00667	.05744	.00308
6	PREPBLKTJA	.00094	-.01318	.00742	-.00549	.00831	.02120
7	S98T002988	7.5573	-.41704	2.9616	-.24267	.77430	46.963
8	S98T002988_D	3.3788	-.43495	2.8330	-.42817	-.00961	33.938
9	S98T002989	3.8227	-.10504	3.6109	-.53799	.63741	20.427
10	S98T002989_D	4.5992	-.49788	3.5465	-.35363	.44891	23.301
11	ICSA	.02394	.01616	-.01769	-.01692	.05817	.00428
12	ICSAB	.02349	.98319	-.02566	-.00874	.04825	.00137
13	CCV	5.0572	5.0648	4.9249	5.1210	4.8859	5.2332
14	CCB	.00234	-.00801	.00269	-.00487	.01135	.00469

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8742	4.9062	.23653	5.1061	4.7799	9.6935
2	ICB	-.00441	-.00004	-.00383	-.00023	-.02389	-.01477
3	LLS	.19186	.01985	.00630	.01929	.39778	.45815
4	ICSA	-.02033	.00172	.00444	.00067	.00033	-.06305
5	ICSAB	-.01800	.00167	.01456	.00115	-.04602	-.05936
6	PREPBLKTJA	.00571	.00008	.00090	-.00061	.00825	-.05593
7	S98T002988	-.05931	.16290	-.18990	.70817	.21950	-2.7240
8	S98T002988_D	.10916	.11544	-.29230	.78851	-.22650	-2.4729
9	S98T002989	.42821	.13905	-.20769	.43042	.03472	-1.2826
10	S98T002989_D	.05334	.13242	-.22348	.40361	-.25848	-2.5958
11	ICSA	-.01451	.00168	.00668	.00112	-.01084	-.05501
12	ICSAB	-.01760	.00170	-.00488	.00042	-.01881	-.06071
13	CCV	4.9096	4.9321	.24943	5.1442	4.7950	9.7166
14	CCB	.00167	.00004	-.00184	-.00044	-.00582	.00723

Analysis Report

Averages

Mon 12-14-98 10:04:54 AM

page 4

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0218	.00841	4.8025	4.8829
2	ICB	-.00064	-.00022	-.00038	-.00029
3	LLS	.09835	-.00019	.01950	.01844
4	ICSA	.00059	.00747	-.00369	-.00404
5	ICSAB	.46367	.00770	.93353	-.00409
6	PREPBLKTJA	.00145	.00025	.00045	.00117
7	S98T002988	.04491	-.00526	.03272	.04045
8	S98T002988_D	.04495	.00747	.05676	.07232
9	S98T002989	.12005	.02946	.00930	.10161
10	S98T002989_D	.05275	.00431	.00585	.04300
11	ICSA	.00211	.00778	-.00392	-.00301
12	ICSAB	.46181	.00778	.92896	-.00288
13	CCV	5.0305	.00809	4.7839	4.8980
14	CCB	.00100	.00000	.00004	.00102

JK

12-14-98

LABCORE Data Entry Template for Worklist# 27268

Analyst: JK 560 Instrument: ICP01 11-8-98

Book# 77830

Method: LA-505-451/161 Rev/Mod C-3

Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysis

Worklist Comment: 11-8-98
ICP BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@ICP-QC	QC		
2 ICB			@ICP-QC	QC		
3 LLS			@ICP-QC	QC		
4 ICSA			@ICP-QC	QC		
5 ICSAB			@ICP-QC	QC		
6 PREPSTDJJA			@ICP-A01	SOLID		
7 PREPBLKTJA			@ICP-A01	SOLID		
8 SERDIL	S98T002990	0 A	@ICP-A01	SOLID		
9 SAMPLE	S98T002990	0 A	@ICP-A01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
10 DUP	S98T002990	0 A	@ICP-A01	SOLID		
11 SPK-PREDIG	S98T002990	0 A	@ICP-A01	SOLID		
12 CCV	<u>spike 24 (10 ppm) 11-8-98</u>		@ICP-QC	QC		
13 CCB			@ICP-QC	QC		
14 SERDIL	S98T002991	0 A	@ICP-A01	SOLID		
15 SAMPLE	S98T002991	0 A	@ICP-A01	SOLID	98000509	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 ,						

Data Entry Comments

uploaded 12-11-98

John Howell

Validated by:

Paul R. Parry

12/14/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

11/20/98 12:21
A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27268

S Type	Sample#	R A	Test	Matrix	Group#	Project		
			BE-A-01	BI-A-01	CA-A-01	CD-A-01	CE-A-01	CO-A-01
			CR-A-01	CU-A-01	FE-A-01	K-A-01	LA-A-01	LI-A-01
			MG-A-01	MN-A-01	MO-A-01	NA-A-01	ND-A-01	NI-A-01
			P-A-01	PB-A-01	S-A-01	SB-A-01	SE-A-01	SI-A-01
			SM-A-01	SR-A-01	TI-A-01	TL-A-01	U-A-01	V-A-01
			ZN-A-01	ZR-A-01				
16 DUP	S98T002991	0 A	@ICP-A01	SOLID				
17 SPK-PREDIG	S98T002991	0 A	@ICP-A01	SOLID				
18 SAMPLE	S98T002992	0 A	@ICP-A01	SOLID	98000509	BY-105 (2)		
			Analytes Requested:					
			AG-A-01	AL-A-01	AS-A-01	B-A-01	BA-A-01	
			BE-A-01	BI-A-01	CA-A-01	CD-A-01	CE-A-01	CO-A-01
			CR-A-01	CU-A-01	FE-A-01	K-A-01	LA-A-01	LI-A-01
			MG-A-01	MN-A-01	MO-A-01	NA-A-01	ND-A-01	NI-A-01
			P-A-01	PB-A-01	S-A-01	SB-A-01	SE-A-01	SI-A-01
			SM-A-01	SR-A-01	TI-A-01	TL-A-01	U-A-01	V-A-01
			ZN-A-01	ZR-A-01				
19 DUP	S98T002992	0 A	@ICP-A01	SOLID				
20 CCV			@ICP-QC	QC				
21 CCB			@ICP-QC	QC				
22 SERDIL	S98T002993	0 A	@ICP-A01	SOLID				
23 SAMPLE	S98T002993	0 A	@ICP-A01	SOLID	98000509	BY-105 (2)		
			Analytes Requested:					
			AG-A-01	AL-A-01	AS-A-01	B-A-01	BA-A-01	
			BE-A-01	BI-A-01	CA-A-01	CD-A-01	CE-A-01	CO-A-01
			CR-A-01	CU-A-01	FE-A-01	K-A-01	LA-A-01	LI-A-01
			MG-A-01	MN-A-01	MO-A-01	NA-A-01	ND-A-01	NI-A-01
			P-A-01	PB-A-01	S-A-01	SB-A-01	SE-A-01	SI-A-01
			SM-A-01	SR-A-01	TI-A-01	TL-A-01	U-A-01	V-A-01
			ZN-A-01	ZR-A-01				
24 DUP	S98T002993	0 A	@ICP-A01	SOLID				
25 SPK-PREDIG	S98T002993	0 A	@ICP-A01	SOLID				
26 ICSA			@ICP-QC	QC				
27 ICSAB			@ICP-QC	QC				

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

11/20/98 12:21
A-0004-1

Page: 3

LABCORE Data Entry Template for Worklist# 27268

S Type	Sample#	R A	Test	Matrix	Group#	Project
28	CCV		ICP-QC	QC		
29	CCB		ICP-QC	QC		

Final page for worklist # 27268

JK [Signature] 12-10-98

Analyst Signature	Date
Prep [Signature]	12-10-98
Prep [Signature]	12-10-98
5987001990-L	12-10-98
5987001990	12-10-98
5987001990-D	12-10-98
5987001990-S	12-10-98
5987001990-X	12-10-98
5987001990-QX	12-10-98
5987001991-L	12-10-98
5987001991	12-10-98
5987001991-D	12-10-98
5987001991-S	12-10-98
5987001991-X	12-10-98
5987001991-QX	12-10-98
5987001992	12-10-98
5987001992	12-10-98
5987001993-L	12-10-98
5987001993	12-10-98
5987001993-D	12-10-98
5987001993-S	12-10-98
5987001993-X	12-10-98
5987001993-QX	12-10-98

Analyst Signature	Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

$$Cr = 3.1954 \frac{\mu g}{ml} \times \frac{1}{5.156 \frac{g}{L} \times \frac{1L}{1000ml}} = 6.20e + 2 \frac{\mu g}{g}$$

$$Pre-dig Spk: \frac{7.2456 - 3.1954}{5} \times 100 = 81.0\%$$

Post Spk:

$$Al = \frac{\left(\frac{524.41}{40}\right) - \left(\frac{135.46}{40}\right)}{10} \times 100 = 97.2\%$$

$$Fe = \frac{\left(\frac{396.33}{40}\right) - \left(\frac{10.920}{40}\right)}{10} \times 100 = 96.35\%$$

$$Na = \frac{\left(\frac{1335.4}{40}\right) - \left(\frac{940.84}{40}\right)}{10} \times 100 = 98.6\%$$

$$P = \frac{\left(\frac{402.66}{40}\right) - \left(\frac{5.6564}{40}\right)}{10} \times 100 = 99.25\%$$

$$S = \frac{\left(\frac{406.76}{40}\right) - \left(\frac{10.434}{40}\right)}{10} \times 100 = 99.1\%$$

$$Si = \frac{\left(\frac{400.01}{40}\right) - \left(\frac{1.0523}{40}\right)}{10} \times 100 = 99.7\%$$

S98T002991

Post Spk:

$$Al = \frac{\left(\frac{562.39}{40}\right) - \left(\frac{188.09}{40}\right)}{10} \times 100 = 93.6\%$$

$$Fe = \frac{\left(\frac{385.50}{40}\right) - \left(\frac{7.7795}{40}\right)}{10} \times 100 = 94.4\%$$

$$Na = \frac{\left(\frac{1214.6}{40}\right) - \left(\frac{877.01}{40}\right)}{10} \times 100 = 84.4\%$$

Post Spike:

$$P = \frac{\left(\frac{404.41}{40}\right) - \left(\frac{17.036}{40}\right)}{10} \times 100 = 96.8\%$$

$$Si = \frac{\left(\frac{399.83}{40}\right) - \left(\frac{7.0248}{40}\right)}{10} \times 100 = 98.2\%$$

S987002993

$$Al = 91.476 \frac{\mu g}{ml} \times \frac{1}{5.014 \frac{g}{L} \times \frac{1L}{1000ml}} = 1.82e+4 \frac{\mu g}{g}$$

$$\text{Pre-dig Spk: } Al = \frac{96.856 - 91.476}{5} \times 100 = 102.2\%$$

Post Spk:

$$Na = \frac{\left(\frac{1340.2}{40}\right) - \left(\frac{978.63}{40}\right)}{10} \times 100 = 90.4\%$$

Analysis Report

Summary

Thu 12-10-98 03:33:59 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981210B	ICP2	12/10/98	13:04	DKS	Q	CONC
2	ICB	981210B	ICP2	12/10/98	13:08	DKS	Q	CONC
3	LLS	981210B	ICP2	12/10/98	13:11	DKS	Q	CONC
4	ICSA	981210B	ICP2	12/10/98	13:14	DKS	Q	CONC
5	ICSAB	981210B	ICP2	12/10/98	13:17	DKS	Q	CONC
6	PREPSTD TJ A	981210B	ICP2	12/10/98	13:20	DKS	Q	CONC
7	PREPBLKTJA	981210B	ICP2	12/10/98	13:23	DKS	Q	CONC
8	S98T002990_L	981210B	ICP2	12/10/98	13:27	DKS	S	CONC
9	S98T002990_	981210B	ICP2	12/10/98	13:30	DKS	S	CONC
10	S98T002990_D	981210B	ICP2	12/10/98	13:33	DKS	S	CONC
11	S98T002990_S	981210B	ICP2	12/10/98	13:36	DKS	S	CONC
12	S98T002990_X	981210B	ICP2	12/10/98	13:39	DKS	S	CONC
13	S98T002990_AX	981210B	ICP2	12/10/98	13:42	DKS	S	CONC
14	CCV	981210B	ICP2	12/10/98	13:52	DKS	Q	CONC
15	CCB	981210B	ICP2	12/10/98	13:55	DKS	Q	CONC
16	S98T002991_L	981210B	ICP2	12/10/98	13:58	DKS	S	CONC
17	S98T002991_	981210B	ICP2	12/10/98	14:01	DKS	S	CONC
18	S98T002991_D	981210B	ICP2	12/10/98	14:05	DKS	S	CONC
19	S98T002991_S	981210B	ICP2	12/10/98	14:07	DKS	S	CONC
20	S98T002991_X	981210B	ICP2	12/10/98	14:12	DKS	S	CONC
21	S98T002991_AX	981210B	ICP2	12/10/98	14:15	DKS	S	CONC
22	S98T002992_	981210B	ICP2	12/10/98	14:23	DKS	S	CONC
23	S98T002992_D	981210B	ICP2	12/10/98	14:26	DKS	S	CONC
24	CCV	981210B	ICP2	12/10/98	14:30	DKS	Q	CONC
25	CCB	981210B	ICP2	12/10/98	14:33	DKS	Q	CONC
26	S98T002993_L	981210B	ICP2	12/10/98	14:39	DKS	S	CONC
27	S98T002993_	981210B	ICP2	12/10/98	14:42	DKS	S	CONC
28	S98T002993_D	981210B	ICP2	12/10/98	14:45	DKS	S	CONC
29	S98T002993_S	981210B	ICP2	12/10/98	14:48	DKS	S	CONC
30	S98T002993_X	981210B	ICP2	12/10/98	14:53	DKS	S	CONC
31	S98T002993_AX	981210B	ICP2	12/10/98	14:56	DKS	S	CONC
32	ICSA	981210B	ICP2	12/10/98	15:21	DKS	Q	CONC
33	ICSAB	981210B	ICP2	12/10/98	15:24	DKS	Q	CONC
34	CCV	981210B	ICP2	12/10/98	15:27	DKS	Q	CONC
35	CCB	981210B	ICP2	12/10/98	15:30	DKS	Q	CONC

JK 6/15/98

12-10-98

Wok/15/ # 27268

84-105

5887001990

5887001991

5887001992

5887001993

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1943 TO 1948.

Analysis Report

Averages

Thu 12-10-98 03:33:59 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9878	4.7885	4.9932	4.9982	4.7869	5.0345
2	ICB	-.00061	-.00810	-.00758	.00565	.00000	.00031
3	LLS	.01573	.10265	.18233	.09969	.09901	.01044
4	ICSA	.00271	237.09	.06388	.00122	.00030	.00041
5	ICSAB	.95399	240.24	.02716	-.00496	.45974	.47380
6	PREPSTD TJ A	.92307	4.6913	4.6965	5.0223	4.7639	4.9113
7	PREPBLKTJ A	-.00070	Q.09141	-.01090	Q.46251	.00069	-.00004
8	S98T002990_L	.05173	138.33	-.11197	.38066	.08016	.00447
9	S98T002990	.06702	134.55	.01081	.28618	.07772	.00211
10	S98T002990_D	.06743	100.61	-.02508	.26902	.09401	.00232
11	S98T002990_S	.95826	192.53	4.7467	4.6977	4.4236	4.5609
12	S98T002990_X	.05810	135.46	-.05764	.43724	.07811	.01443
13	S98T002990_AX	407.32	524.41	404.83	391.28	401.27	398.62
14	CCV	4.9923	4.8873	5.0507	5.0847	4.9070	5.1123
15	CCB	.00287	.00828	.00818	.00612	.00030	.00031
16	S98T002991_L	.05797	188.88	.04530	.37927	.05655	.00718
17	S98T002991	.06066	184.84	-.00877	.30944	.05628	.00174
18	S98T002991_D	.06450	159.84	.02171	.40629	.05445	.00158
19	S98T002991_S	1.0057	184.40	4.8279	4.7643	4.6003	4.6948
20	S98T002991_X	-.02445	188.09	.08077	.39825	.05169	.01266
21	S98T002991_AX	397.72	562.39	395.61	381.12	389.78	389.31
22	S98T002992	.07437	198.17	.03294	.13987	.06518	.00248
23	S98T002992_D	.06069	171.46	-.00756	.24023	.07328	.00212
24	CCV	5.0083	4.8605	5.0360	5.0504	4.8721	5.0934
25	CCB	-.00021	.00185	-.00354	.00706	.00013	.00031
26	S98T002993_L	.04933	92.120	-.02912	.31270	.03243	.00630
27	S98T002993	.06380	91.476	-.02587	.25773	.03152	.00178
28	S98T002993_D	.06464	94.955	-.03730	.39859	.03214	.00178
29	S98T002993_S	.28640	96.856	4.8015	4.7444	4.4434	4.4995
30	S98T002993_X	.01027	91.956	.11042	.36039	.03631	.01085
31	S98T002993_AX	239.06	468.90	393.40	379.09	384.89	386.72
32	ICSA	.01929	236.32	.03571	-.00597	.00030	.00022
33	ICSAB	.95232	239.12	.05099	-.00410	.45459	.46669
34	CCV	4.9560	4.8093	4.9559	4.9875	4.7874	5.0155
35	CCB	.00531	-.00572	.00382	.00236	.00000	.00027

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0125	5.3348	5.1276	4.8757	5.0134	5.0495
2	ICB	-.04776	.00006	-.00139	-.00986	-.00157	-.00105
3	LLS	.18560	.23852	.00875	.19323	.04011	.02068
4	ICSA	-.04117	258.20	.00399	.00979	.00148	-.00159
5	ICSAB	-.02954	260.33	.94508	.00625	.46612	.47270
6	PREPSTD TJ A	4.5276	5.0963	4.6378	4.7772	4.6506	4.6960
7	PREPBLKTJ A	-.03717	Q.15939	-.00063	.00036	-.00024	.00406
8	S98T002990_L	-.38514	10.193	-.00529	-.15578	-.04260	3.2335
9	S98T002990	-.06088	10.037	.01915	.02293	.01725	3.1954
10	S98T002990_D	-.19559	37.972	.01500	.02982	.01016	2.5082
11	S98T002990_S	4.3415	12.183	4.5819	4.4631	4.4762	7.2456
12	S98T002990_X	-.58794	10.194	.00398	.00651	-.00457	3.2660
13	S98T002990_AX	412.55	428.59	403.13	397.37	401.91	410.14
14	CCV	5.1135	5.2673	5.1170	5.0021	5.0110	5.0333

Analysis Report

Averages

Thu 12-10-98 03:33:59 PM

page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
15	CCB	-.03648	.00880	.00096	.01501	-.00009	.00115
16	S98T002991_L	-.58161	7.8807	.00633	-.05739	-.01123	2.9257
17	S98T002991	.04885	7.6096	.02408	-.00113	.02517	2.8459
18	S98T002991_D	-.06676	4.3339	.01551	-.02409	.01566	2.4040
19	S98T002991_S	4.5641	10.578	4.7048	4.6643	4.6438	7.2513
20	S98T002991_X	-.92827	7.6776	-.02722	-.52509	.02917	2.9625
21	S98T002991_AX	404.17	420.10	396.22	386.66	395.86	402.47
22	S98T002992	-.07269	7.3714	.03037	.02256	.02989	1.8208
23	S98T002992_D	-.05610	13.510	.02203	-.03413	.01355	1.8090
24	CCV	5.0869	5.3031	5.1406	4.9665	5.0417	5.0569
25	CCB	-.04184	.00449	-.00017	.00045	.00090	.00073
26	S98T002993_L	-.85630	5.1939	-.00161	-.00965	-.02700	2.1409
27	S98T002993	-.09914	5.0458	.00708	-.00548	.00926	2.1150
28	S98T002993_D	-.11953	11.438	.00518	-.00926	.00479	2.3750
29	S98T002993_S	4.4420	10.702	4.6139	4.4366	4.5337	6.6565
30	S98T002993_X	-1.7060	5.3120	-.02765	-.23093	-.15701	2.1445
31	S98T002993_AX	399.74	417.30	393.98	382.55	394.49	400.50
32	ICSA	-.05495	251.81	.00390	.00804	.00049	-.00253
33	ICSAB	-.04598	255.02	.92608	.00393	.45424	.46158
34	CCV	5.0166	5.2524	5.0620	4.8895	4.9667	4.9905
35	CCB	-.04189	.00053	-.00081	-.00650	-.00033	.00010

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.5518	.00893	4.8002	4.5482	4.9054	4.6789
2	ICB	-.00076	.00178	-.00183	-.22189	-.00123	.00124
3	LLS	.01725	.00161	.10279	.44478	.09945	.02158
4	ICSA	.00346	.00134	91.896	.36275	-.00335	.00365
5	ICSAB	.45081	.00423	92.828	.12046	-.00462	.93999
6	PREPSTDTJA	4.3810	.00695	4.6422	4.3344	4.7656	4.6669
7	PREPBLKTJA	Q.05254	.00071	Q.05128	-.07940	-.00001	.00072
8	S98T002990_L	.02627	.05744	11.125	3.5364	-.01287	.04163
9	S98T002990	.04341	.02115	10.792	7.3495	.00970	.01413
10	S98T002990_D	.03982	.04959	15.269	5.5361	.00704	.01206
11	S98T002990_S	4.1597	.02054	11.959	9.9881	4.4536	4.2657
12	S98T002990_X	.05152	.09855	10.930	6.7293	.01698	.07494
13	S98T002990_AX	365.57	.61100	396.33	395.76	396.19	382.01
14	CCV	4.6842	.00760	4.8261	4.8294	5.0241	4.9647
15	CCB	.00088	.00150	.00050	.24128	.00184	.00135
16	S98T002991_L	.04222	.02946	7.8824	4.4997	-.00660	.02496
17	S98T002991	.03850	.01432	7.6262	5.8506	.00385	.00870
18	S98T002991_D	.02923	.00874	5.1810	5.0412	-.00088	.00537
19	S98T002991_S	4.3614	.01689	9.6019	10.788	4.6366	4.5338
20	S98T002991_X	.06472	.06792	7.7795	-.7.7943	-.06694	.06664
21	S98T002991_AX	355.98	.60803	385.50	375.86	385.37	367.72
22	S98T002992	.06944	.01423	11.825	3.4350	.00891	.01247
23	S98T002992_D	.06316	.02168	13.754	3.5175	.00073	.01373
24	CCV	4.6400	.00804	4.8329	4.8004	4.9927	4.8127
25	CCB	.00039	.00202	-.00041	.12265	.00051	.00166
26	S98T002993_L	.00530	.06271	5.2539	6.2290	.00128	.05623
27	S98T002993	.02244	.01438	5.1772	4.9624	.00307	.01833
28	S98T002993_D	.01418	.01885	4.8118	4.5256	.00088	.01625
29	S98T002993_S	4.1408	.02275	11.286	8.7394	4.4004	4.2511

Analysis Report

Averages

Thu 12-10-98 03:33:59 PM

page 4

#	Sample Name	Cu	Eu	Fe	K	La	Li
30	S98T002993_X	.00028	.12077	5.1921	11.528	-.00408	.10830
31	S98T002993_AX	351.30	.63372	381.09	379.46	380.99	364.97
32	ICSA	.00303	.01000	90.183	.09299	-.00345	.00302
33	ICSAB	.44721	.01010	91.266	.05845	-.00383	.94991
34	CCV	4.5788	.00870	4.7477	4.8828	4.9191	4.8064
35	CCB	-.00011	.00212	-.00136	-.09203	-.00123	.00135

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.0132	4.9560	5.0590	4.6598	4.7293	5.0281
2	ICB	-.01721	.00064	.00183	-.03166	-.00309	-.00477
3	LLS	.19600	.02085	.10446	.17677	.19490	.04093
4	ICSA	243.35	-.00396	-.01209	181.85	-.00072	-.00120
5	ICSAB	246.14	.44672	-.01524	184.80	-.00004	.91770
6	PREPSTDITJA	4.6528	4.6457	4.7918	5.8654	4.6464	4.6234
7	PREPBLKTJA	.03953	.00138	.00457	Q.74508	.00069	.00754
8	S98T002990_L	1.3184	.37006	.06999	976.66	-.04894	.35484
9	S98T002990	1.4856	.34763	.04957	931.23	.01116	.32999
10	S98T002990_D	1.4387	.31388	.03572	912.93	.00926	.30265
11	S98T002990_S	5.9702	4.5638	4.5729	857.87	4.3108	4.7401
12	S98T002990_X	.96400	.38518	.07396	940.84	.05808	.02686
13	S98T002990_AX	410.76	404.63	400.57	1335.4	383.02	405.82
14	CCV	5.1256	4.9536	5.0752	4.8925	4.8557	5.0203
15	CCB	.00574	.00057	.00247	.01078	.00131	-.00152
16	S98T002991_L	.79522	.26310	.06628	887.53	-.00710	.41767
17	S98T002991	.93357	.24901	.04828	857.53	.01843	.40394
18	S98T002991_D	.98607	.16665	.04366	964.07	.01236	.35612
19	S98T002991_S	7.2456	4.7101	4.7224	1020.8	4.5064	5.0255
20	S98T002991_X	.31131	.28013	.15181	877.01	.08686	.14806
21	S98T002991_AX	399.86	395.89	393.01	1214.6	371.22	397.59
22	S98T002992	1.6674	.34954	.02203	990.37	.01738	.26544
23	S98T002992_D	4.2069	.42738	.02199	964.53	.00811	.25793
24	CCV	5.1057	4.9646	5.0895	4.7795	4.8185	5.0218
25	CCB	-.00278	.00078	.00461	-.01584	.00193	-.00432
26	S98T002993_L	.78959	.15720	.06365	969.67	-.07672	.24095
27	S98T002993	.85874	.13872	.04350	962.83	.00518	.29992
28	S98T002993_D	.69135	.13735	.04414	1013.0	-.00185	.30509
29	S98T002993_S	5.4381	4.5328	4.5475	873.23	4.2545	4.7856
30	S98T002993_X	.57080	.17982	.10733	978.63	-.01898	.13809
31	S98T002993_AX	397.87	393.85	392.03	1340.2	366.75	396.79
32	ICSA	242.48	-.00440	-.01436	184.09	-.00101	-.01266
33	ICSAB	245.08	.43584	-.01555	185.50	-.00253	.90032
34	CCV	5.0549	4.8916	5.0204	4.7401	4.7461	4.9559
35	CCB	-.01459	.00075	.00363	-.01222	-.00081	-.00247

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0103	5.1092	4.9246	5.1119	4.8376	5.2019
2	ICB	.02353	.00447	.01545	.00375	.03847	.00909
3	LLS	.41409	.20655	.21801	.10294	.22442	.15196
4	ICSA	.05622	.04580	-.03178	.01577	.07447	.00311
5	ICSAB	.05474	1.0329	-.00823	.00142	.08467	.00225

Analysis Report

Averages

Thu 12-10-98 03:33:59 PM

page 5

#	Sample Name	P	Pb	S	Sb	Se	Si
6	PREPSTD TJ A	4.5884	4.5812	4.4444	4.6217	4.4914	7.5034
7	PREPBLKTJA	.04527	.00111	Q.09357	.00984	.04481	Q.75822
8	S98T002990_L	5.8389	.11033	10.462	.11428	.88867	.59625
9	S98T002990	5.4055	.37900	10.323	-.01190	.24818	.26066
10	S98T002990_D	22.065	.34197	6.4557	.01576	.27894	3.0877
11	S98T002990_S	13.240	4.8512	10.791	4.3838	4.6558	11.556
12	S98T002990_X	5.6564	.56663	10.434	.34919	1.6933	1.0523
13	S98T002990_AX	402.66	412.16	406.76	404.50	395.80	400.01
14	CCV	5.0866	5.0907	4.9662	5.1312	4.9015	5.2056
15	CCB	.02448	.00390	.01440	.00617	.03478	.01025
16	S98T002991_L	16.435	.45122	9.4676	.23859	.62467	6.5586
17	S98T002991	16.056	.33711	9.2840	.00581	.26216	6.3592
18	S98T002991_D	13.718	.31476	10.519	.02044	.23411	5.7720
19	S98T002991_S	16.372	5.0016	14.509	4.6516	4.7168	13.253
20	S98T002991_X	17.036	.74590	9.6272	-.24581	1.4816	7.0248
21	S98T002991_AX	404.41	406.17	396.07	397.83	386.37	399.83
22	S98T002992	17.388	.35713	8.1342	.06228	.33118	6.2654
23	S98T002992_D	12.719	.34303	8.2216	-.00808	.24905	5.5775
24	CCV	5.0640	5.1149	4.9764	5.1595	4.8684	5.2092
25	CCB	.01466	.01032	.01009	.00536	.04729	.01357
26	S98T002993_L	8.5784	.13954	6.9781	.07585	1.0038	5.2205
27	S98T002993	8.1545	.24922	6.9131	.02993	.20650	4.9189
28	S98T002993_D	8.3232	.16117	7.6030	.00216	.19625	5.4909
29	S98T002993_S	14.125	4.8066	11.364	4.4134	4.6160	10.824
30	S98T002993_X	9.2719	.18656	6.9594	.21066	2.7131	5.7866
31	S98T002993_AX	392.28	405.76	394.39	395.76	380.32	396.38
32	ICSA	.05864	.01895	-.01536	.00002	.05102	.00168
33	ICSRAB	.05208	.99847	-.02704	.01892	.07729	.00149
34	CCV	4.9920	5.0611	4.8941	5.0812	4.8227	5.1421
35	CCB	.02572	.00406	.01166	.00057	.03804	.01079

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.7431	4.7811	.20005	5.0272	4.7945	9.3765
2	ICB	.01780	.00021	-.00539	.00020	-.00779	.06349
3	LLS	.21296	.02005	.00132	.02040	.39554	.52105
4	ICSA	.00034	.00169	.00651	.00186	.02291	-.01338
5	ICSRAB	-.00370	.00175	.01372	.00234	.01737	-.02223
6	PREPSTD TJ A	4.6811	4.7671	.18607	4.8460	4.4302	9.0918
7	PREPBLKTJA	.00389	.00029	.00105	.00160	.00482	.01829
8	S98T002990_L	.44688	.39646	-.12674	.76292	-.10579	2.4138
9	S98T002990	.09262	.38052	.02363	.75814	.01326	1.1220
10	S98T002990_D	.07362	.32983	.03311	.82595	-.00775	.89084
11	S98T002990_S	4.3616	4.7733	.20765	4.9982	4.2135	9.1444
12	S98T002990_X	.91394	.39718	-.09551	.76740	-.12170	4.1691
13	S98T002990_AX	383.33	392.19	16.835	395.84	393.87	740.32
14	CCV	4.8912	4.8929	.20666	5.1018	4.8428	9.6421
15	CCB	.01791	.00025	.00803	.00071	-.00469	.06176
16	S98T002991_L	.23682	.43935	.00409	1.0562	.01753	1.7529
17	S98T002991	.05267	.42633	.02082	1.0335	-.08139	1.1728
18	S98T002991_D	.02464	.36834	.00592	.46066	-.09174	1.0404
19	S98T002991_S	4.5496	5.0556	.23968	4.9836	4.4291	9.8431
20	S98T002991_X	.65938	.43922	-.39313	1.0591	-.36534	3.3266

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
21	S98T002991_AX	371.64	381.14	16.642	387.61	388.80	720.82
22	S98T002992	.04981	.30750	.02874	1.1883	.00400	.96403
23	S98T002992_D	.05613	.25285	-.01329	1.1771	-.01684	.86328
24	CCV	4.8438	4.8503	.21016	5.0652	4.7913	9.5450
25	CCB	.02164	.00029	-.00018	.00067	.00492	.08000
26	S98T002993_L	.61572	.21423	-.11701	.46112	-.09631	2.7159
27	S98T002993	.09262	.20638	-.00355	.45998	-.01086	.77428
28	S98T002993_D	.04832	.23199	-.00656	.43429	-.05640	.69881
29	S98T002993_S	4.3433	4.6136	.17911	4.9112	4.3697	8.9909
30	S98T002993_X	1.3836	.22273	-.18688	.46527	-.07492	5.2292
31	S98T002993_AX	367.37	377.08	16.287	383.38	383.60	710.55
32	ICSA	-.00242	.00170	.00649	.00162	.00980	-.00147
33	ICSAB	-.00571	.00176	.00153	.00160	.00523	-.00502
34	CCV	4.7746	4.7933	.21714	5.0209	4.7308	9.4281
35	CCB	.02464	.00029	-.00657	.00042	.01879	.07869

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9675	.00925	4.8467	4.8251
2	ICB	.00360	.00114	-.00006	.00315
3	LLS	.10297	.00134	.01992	.02190
4	ICSA	.00569	.00854	-.00350	-.00086
5	ICSAB	.46768	.00839	.94243	-.00253
6	PREPSTDJTA	4.8043	.00718	4.3128	4.6380
7	PREPBLKTJA	.00152	.00046	Q.03642	.00303
8	S98T002990_L	.10937	.03345	.05406	.10024
9	S98T002990	.04078	.01174	.06348	.03771
10	S98T002990_D	.05658	.01216	.06433	.03257
11	S98T002990_S	4.5678	.01366	4.4369	4.4143
12	S98T002990_X	.25093	.06719	.07184	.17751
13	S98T002990_AX	395.09	.68539	384.08	393.97
14	CCV	5.0051	.00852	4.8016	4.8845
15	CCB	.00410	.00107	.00041	.00184
16	S98T002991_L	.07907	.01806	.05804	.05368
17	S98T002991	.04697	.00702	.05431	.03175
18	S98T002991_D	.02412	.00520	.04156	.01953
19	S98T002991_S	4.7093	.01180	4.5503	4.5342
20	S98T002991_X	.19018	.04560	.06379	.16298
21	S98T002991_AX	387.84	.66860	378.92	385.37
22	S98T002992	.05835	.00962	.13374	.03490
23	S98T002992_D	.05232	.01029	.13381	.03789
24	CCV	5.0069	.00871	4.8305	4.8790
25	CCB	.00479	.00145	.00010	.00305
26	S98T002993_L	.16329	.03646	.04307	.11842
27	S98T002993	.04438	.00719	.03737	.02666
28	S98T002993_D	.03037	.00565	.03352	.02309
29	S98T002993_S	4.5953	.01561	4.5280	4.3614
30	S98T002993_X	.32133	.07928	.04438	.22119
31	S98T002993_AX	385.50	.69177	379.56	382.86
32	ICSA	.00392	.00799	-.00382	-.00195
33	ICSAB	.46076	.00815	.92790	-.00116
34	CCV	4.9376	.00868	4.7777	4.8065
35	CCB	.00481	.00137	.00017	.00353

JK
12-10-98

LABCORE Data Entry Template for Worklist# 27442

Analyst: DK Seb Instrument: ICP01 11-11-98Book# 77B30Method: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysisWorklist Comment: DK 11-18-98
ICP BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTDJTA		@ICP-A01	SOLID		
7	PREPBLKTJA		@ICP-A01	SOLID		
8	SERDIL	S98T002604 0 A	@ICP-A01	SOLID		
9	SAMPLE	S98T002604 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
10	DUP	S98T002604 0 A	@ICP-A01	SOLID		
11	SPK-PREDIG	S98T002604 0 A	@ICP-A01	SOLID		
12	CCV	<u>52100 08 (10 ppm)</u>	@ICP-QC	QC		
13	CCB		@ICP-QC	QC		
14	SAMPLE	S98T002605 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 ,						

Data Entry Comments:

uploaded 12-30-98
JK Howell

Validated by:

Sam M. Lang12/30/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27442

S Type	Sample#	R A	Test	Matrix	Group#	Project
			MG-A-01	MN-A-01	MO-A-01	NA-A-01
			P-A-01	PB-A-01	S-A-01	SB-A-01
			SM-A-01	SR-A-01	TI-A-01	TL-A-01
			ZN-A-01	ZR-A-01	U-A-01	V-A-01
15 DUP	S98T002605	0 A	@ICP-A01	SOLID		
16 SAMPLE	S98T002606	0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
	Analytes Requested:					
			AG-A-01	AL-A-01	AS-A-01	B-A-01
			BE-A-01	BI-A-01	CA-A-01	CD-A-01
			CR-A-01	CU-A-01	FE-A-01	K-A-01
			MG-A-01	MN-A-01	MO-A-01	NA-A-01
			P-A-01	PB-A-01	S-A-01	SB-A-01
			SM-A-01	SR-A-01	TI-A-01	TL-A-01
			ZN-A-01	ZR-A-01	U-A-01	V-A-01
17 DUP	S98T002606	0 A	@ICP-A01	SOLID		
18 ICSEA			@ICP-QC	QC		
19 ICSAB			@ICP-QC	QC		
20 CCV			@ICP-QC	QC		
21 CCB			@ICP-QC	QC		

Final page for worklist # 27442

JK *12-11-98*

Analyst Signature	Date
Preps 4/ JTA	12-11-98
Prep 6/ JTA	12-11-98
5987002604-L	12-8-98
5987002604	12-8-98
5987002604-D	12-8-98
5987002604-X	12-8-98
5987002604-X	12-8-98

Analyst Signature	Date
5987002605	12-8-98
5987002605-D	12-8-98
5987002606	12-8-98
5987002606-D	12-8-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

$$C_2 = 0.27087 \frac{\mu\text{g}}{\text{ml}} \times \frac{1}{5.428 \frac{\text{g}}{\text{L}} \times \frac{1\text{L}}{100\text{ml}}} = 4.992 \pm 1 \frac{\mu\text{g}}{\text{g}}$$

Pre-dig Spk:

$$C_2 = \frac{5.0990 - 0.27087}{5} \times 100 = 96.6\%$$

Post Spike:

$$Al = \frac{\left(\frac{489.67}{50}\right) - \left(\frac{14.867}{50}\right)}{10} \times 100 = 95.0\%$$

$$Na = \frac{\left(\frac{1889.8}{50}\right) - \left(\frac{1454.6}{50}\right)}{10} \times 100 = 87.0\%$$

Analysis Report

Summary

Mon 12-21-98 02:30:44 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981221B	ICP2	12/21/98	12:45	DKS	Q	CONC
2	ICB	981221B	ICP2	12/21/98	12:48	DKS	Q	CONC
3	LLS	981221B	ICP2	12/21/98	12:51	DKS	Q	CONC
4	ICSA	981221B	ICP2	12/21/98	12:54	DKS	Q	CONC
5	ICSAB	981221B	ICP2	12/21/98	12:57	DKS	Q	CONC
6	PREPSTD TJJA	981221B	ICP2	12/21/98	13:00	DKS	Q	CONC
7	PREPBLK TJJA	981221B	ICP2	12/21/98	13:03	DKS	Q	CONC
8	S98T002604_L	981221B	ICP2	12/21/98	13:07	DKS	S	CONC
9	S98T002604	981221B	ICP2	12/21/98	13:10	DKS	S	CONC
10	S98T002604_D	981221B	ICP2	12/21/98	13:14	DKS	S	CONC
11	S98T002604_S	981221B	ICP2	12/21/98	13:17	DKS	S	CONC
12	S98T002604_X	981221B	ICP2	12/21/98	13:21	DKS	S	CONC
13	S98T002604_AX	981221B	ICP2	12/21/98	13:25	DKS	S	CONC
14	CCV	981221B	ICP2	12/21/98	13:51	DKS	Q	CONC
15	CCB	981221B	ICP2	12/21/98	13:57	DKS	Q	CONC
16	S98T002605	981221B	ICP2	12/21/98	14:00	DKS	S	CONC
17	S98T002605_D	981221B	ICP2	12/21/98	14:03	DKS	S	CONC
18	S98T002606	981221B	ICP2	12/21/98	14:07	DKS	S	CONC
19	S98T002606_D	981221B	ICP2	12/21/98	14:10	DKS	S	CONC
20	ICSA	981221B	ICP2	12/21/98	14:16	DKS	Q	CONC
21	ICSAB	981221B	ICP2	12/21/98	14:21	DKS	Q	CONC
22	CCV	981221B	ICP2	12/21/98	14:24	DKS	Q	CONC
23	CCB	981221B	ICP2	12/21/98	14:27	DKS	Q	CONC

JK LJP

12-21-98

Work list # 27442

BY 105

598T002604

598T002605

598T002606

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 1952 TO 1955.

Analysis Report

Averages

Mon 12-21-98 02:30:44 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9803	4.8877	5.0661	5.1139	4.8861	5.0827
2	ICB	.00651	-.00098	-.00205	-.00277	.00000	-.00008
3	LLS	Q.02682	.10629	.19670	.09735	.09860	.01029
4	ICSA	.00787	243.06	-.01048	-.00916	.00035	.00004
5	ICSAB	.94898	241.52	.02046	-.01050	.45976	.47304
6	PREPSTD TJ A	Q.14596	4.8766	4.9318	5.2760	4.8442	4.8602
7	PREPBLKTJ A	.00114	Q.09690	.00348	Q.48793	.00039	.00008
8	S98T002604_L	.10237	15.247	.06897	.67205	.00026	.00000
9	S98T002604	.10534	14.819	.00058	.70671	.00088	.00000
10	S98T002604_D	.10249	2.9703	-.01136	.58390	.00089	.00021
11	S98T002604_S	.25646	7.9042	5.1399	5.6174	4.7390	4.8825
12	S98T002604_X	.03117	14.867	.61227	.51045	-.00363	.00213
13	S98T002604_AX	306.53	489.67	495.23	483.94	483.87	483.76
14	CCV	4.9511	4.8445	5.0177	5.0805	4.8278	5.0516
15	CCB	.00835	-.00192	-.00095	-.00231	-.00008	.00004
16	S98T002605	.09815	18.933	.03211	.70901	.00076	.00000
17	S98T002605_D	.10519	16.617	.00826	.64183	.00078	.00000
18	S98T002606	.11596	93.001	-.00948	.72893	.01656	.00124
19	S98T002606_D	.11245	99.863	.00988	.67126	.01795	.00146
20	ICSA	.00708	239.75	-.00836	-.00850	.00029	.00008
21	ICSAB	.94497	240.24	.00631	-.00851	.45458	.46638
22	CCV	4.9631	4.8454	5.0289	5.0555	4.8123	5.0435
23	CCB	.00319	-.00636	-.00598	-.00416	-.00009	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0592	5.1600	5.1146	5.0386	4.9726	5.0106
2	ICB	-.00900	.00112	.00027	.00035	.00000	.00074
3	LLS	.18327	.23441	.01044	.20413	.04056	.02024
4	ICSA	-.03792	251.88	.00387	.01083	.00396	-.00024
5	ICSAB	-.02347	253.40	.94161	.00469	.45997	.46568
6	PREPSTD TJ A	4.7046	5.3244	4.7998	4.9326	4.7417	4.7827
7	PREPBLKTJ A	-.01980	Q.17686	.00068	.00455	.00091	.00186
8	S98T002604_L	-.56478	.26061	.00670	-.08083	-.01437	.25516
9	S98T002604	-.05791	.33733	.00836	.02962	.00177	.27087
10	S98T002604_D	-.01170	.82851	.00727	.02142	.00469	.16275
11	S98T002604_S	5.0424	5.6460	5.0399	4.9210	4.8959	5.0990
12	S98T002604_X	-.69039	.05376	-.01745	-.28943	-.06261	.20672
13	S98T002604_AX	494.87	501.96	491.40	485.41	487.35	494.58
14	CCV	5.0421	5.1591	5.0873	4.9912	4.9518	4.9914
15	CCB	-.02952	-.00073	.00035	-.00286	.00036	-.00061
16	S98T002605	-.16005	.24075	.00343	-.01758	.00588	.55785
17	S98T002605_D	-.12941	.94541	.00568	.00047	.00591	.61008
18	S98T002606	.01614	1.1062	.04731	.03706	.03037	2.4012
19	S98T002606_D	.02244	1.1969	.05378	-.01165	.03276	2.8541
20	ICSA	-.03485	251.19	.00423	.00328	.00058	-.00176
21	ICSAB	-.03982	250.94	.93208	.01306	.45380	.46399
22	CCV	5.0273	5.1755	5.0938	4.9767	4.9574	5.0031
23	CCB	-.01256	-.00283	-.00064	-.00518	.00035	-.00048

Analysis Report

Averages

Mon 12-21-98 02:30:44 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7283	.00829	4.8166	5.0125	5.0592	5.0030
2	ICB	.00018	.00027	.00060	-.05641	.00072	.00021
3	LLS	.01702	.00056	.09831	.38686	.10212	.02102
4	ICSA	.00289	.03451	92.726	.22130	-.00319	.00281
5	ICSAB	.45524	.03307	92.661	-.01729	-.00424	.98393
6	PREPSTD TJ A	4.5541	.00873	4.7621	4.8801	4.8879	4.8999
7	PREPBLKT J A	.00422	-.00040	.03085	.10593	.00101	-.00010
8	S98T002604_L	-.00100	.01325	.16640	-3.1378	.01415	.02709
9	S98T002604	.00643	.00284	.18642	1.2326	.00465	.00921
10	S98T002604_D	.00323	.00298	.16363	.80150	.00513	.00596
11	S98T002604_S	4.5738	.01513	5.5034	6.0002	4.8503	4.7451
12	S98T002604_X	-.04755	.08262	.12678	-5.9991	-.01936	.07035
13	S98T002604_AX	448.99	.84948	469.63	486.76	482.61	473.33
14	CCV	4.6661	.00843	4.7913	4.8824	4.9992	4.9373
15	CCB	-.00017	.00093	-.00129	-.17766	.00018	.00097
16	S98T002605	.00363	.00439	.11456	1.2185	-.00013	.00541
17	S98T002605_D	.00326	.00647	.13343	2.7721	.00135	.00649
18	S98T002606	.01241	.00512	3.3708	2.5384	.01190	.09594
19	S98T002606_D	.00668	.00891	3.8898	3.6677	.00720	.11003
20	ICSA	.00203	.02928	91.938	-.01188	-.00397	.00325
21	ICSAB	.45251	.03412	91.723	.18394	-.00272	.98437
22	CCV	4.6569	.00932	4.7823	5.0871	4.9930	4.9185
23	CCB	-.00065	.00064	-.00089	-.11080	-.00103	.00032

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1424	4.9122	5.0520	4.9890	4.9104	5.0095
2	ICB	.00314	-.00002	.00077	.01102	.00230	.00634
3	LLS	.20558	.01993	.10206	.21754	.19923	.04660
4	ICSA	249.95	-.00558	-.01333	193.90	.00317	-.00183
5	ICSAB	248.00	.43994	-.01386	192.26	-.00068	.91686
6	PREPSTD TJ A	4.8798	4.7378	4.8583	6.2569	4.7791	4.7718
7	PREPBLKT J A	.04427	.00029	.00045	0.80784	.00066	.00438
8	S98T002604_L	.14275	.00230	.02745	1480.5	-.01076	.15871
9	S98T002604	.06966	.00189	.00467	1422.0	.01217	.02462
10	S98T002604_D	.07857	.00162	-.00153	1373.0	.00137	.02572
11	S98T002604_S	4.9876	4.7696	5.0251	1305.1	4.7149	4.8889
12	S98T002604_X	.13310	.01982	-.02907	1454.6	-.13577	.08129
13	S98T002604_AX	498.82	488.96	487.50	1889.8	466.69	493.50
14	CCV	5.0788	4.8774	5.0510	4.8946	4.8412	4.9771
15	CCB	.00165	.00015	-.00200	-.00661	.00172	.00711
16	S98T002605	.00632	.00463	.01093	1415.7	.01102	.05095
17	S98T002605_D	.03328	.00452	.01873	1418.9	.00836	.03497
18	S98T002606	.16120	.13863	.02485	1449.4	.01227	.46430
19	S98T002606_D	.15065	.15220	.02659	1487.3	.00639	.51848
20	ICSA	246.99	-.00529	-.00990	189.63	.00090	.00094
21	ICSAB	246.80	.43460	-.01306	191.31	-.00036	.90496
22	CCV	5.0864	4.8910	5.0382	4.8862	4.8372	4.9728
23	CCB	-.00903	.00000	.00125	-.01063	-.00026	.00131

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.0247	5.0616	4.9021	5.1042	4.8798	5.1370
2	ICB	.00226	.01545	-.01084	.00512	.01518	.00289
3	LLS	.40781	.20400	.18291	.09858	.19858	Q.23979
4	ICSA	.03039	.02114	-.03720	.01154	.04848	-.00611
5	ICSAB	.03166	.99670	-.03086	.00697	.06339	-.00118
6	PREPSTDITJA	4.7660	4.7644	4.5986	4.7651	4.5999	1.6850
7	PREPBLKTJA	.02713	.01181	Q.06294	.00634	.01099	Q.42531
8	S98T002604_L	1.1570	.05350	.04222	-.00092	.60677	.67644
9	S98T002604	.84967	.06304	.33954	.01457	.12885	.56751
10	S98T002604_D	-.33943	.00432	.31151	.02307	.09333	.51162
11	S98T002604_S	4.6258	5.0940	5.1571	5.0144	4.8598	5.5093
12	S98T002604_X	1.2250	-.64694	-.04741	-.25312	1.2228	1.0219
13	S98T002604_AX	487.16	501.45	481.45	494.50	479.92	482.17
14	CCV	5.0114	5.0470	4.8530	5.0956	4.8473	5.1041
15	CCB	.01175	-.00754	-.01242	.00289	.02441	.00394
16	S98T002605	1.0441	.03482	1.6169	-.00287	.18840	.54294
17	S98T002605_D	1.0980	.06984	1.6529	-.01481	.05328	.44816
18	S98T002606	19.480	.27309	26.853	.01814	.12721	1.8171
19	S98T002606_D	17.542	.23381	32.561	.00093	.12827	1.6302
20	ICSA	.03951	.02372	-.03522	.00061	.08453	-.00379
21	ICSAB	.03547	.99731	-.02918	.00105	.05905	-.00100
22	CCV	5.0229	5.0945	4.8786	5.0778	4.8647	5.1178
23	CCB	.00873	-.00096	-.01135	-.00049	.02036	.00371

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.9243	4.9273	.21505	5.0918	4.9071	9.6867
2	ICB	.00362	.00000	-.00040	.00000	.00853	.00075
3	LLS	.20339	.02000	.00119	.02001	.39885	.47338
4	ICSA	-.00892	.00195	.00534	.00173	-.00131	.05681
5	ICSAB	-.01166	.00194	.00709	.00107	-.02421	.05640
6	PREPSTDITJA	4.8050	4.8799	.20730	4.9029	4.5965	9.3157
7	PREPBLKTJA	-.00841	.00012	-.00034	.00087	.00981	-.02641
8	S98T002604_L	.18150	.00636	-.06960	-.00571	.10213	.45202
9	S98T002604	.03190	.00383	.01385	.00116	.01467	.06117
10	S98T002604_D	.03756	.00274	.00766	.00006	.02153	.03532
11	S98T002604_S	4.7754	4.8324	.20844	4.9435	4.8187	9.3610
12	S98T002604_X	.96050	.01254	-.48239	-.02398	.10011	3.1770
13	S98T002604_AX	467.25	476.74	21.132	478.06	486.81	904.14
14	CCV	4.8643	4.8651	.20628	5.0459	4.8003	9.5777
15	CCB	.01230	.00012	-.00335	-.00001	-.00375	.04045
16	S98T002605	.05915	.00895	-.02767	.00209	-.04296	.15149
17	S98T002605_D	.07504	.00871	-.01837	.00433	.07112	.18239
18	S98T002606	.04334	.54961	.06123	.01713	.01435	1.3762
19	S98T002606_D	.07700	.62528	.03843	.02352	.00583	1.6827
20	ICSA	-.00613	.00193	.00512	.00150	-.04012	.07593
21	ICSAB	-.00776	.00198	.00661	.00150	-.01858	.07932
22	CCV	4.8335	4.8616	.22463	5.0394	4.8232	9.5597
23	CCB	.00619	.00004	-.00858	-.00025	-.00272	.01880

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9906	.00803	4.7344	4.8625
2	ICB	.00067	.00000	-.00015	.00062
3	LLS	.09952	.00003	.01983	.01989
4	ICSA	.00081	.00734	-.00417	-.00360
5	ICSAB	.46341	.00773	.92380	-.00294
6	PREPSTD TJA	4.8971	.00709	4.4602	4.6799
7	PREPBLKTJA	-.00182	-.00069	.00759	.00065
8	S98T002604_L	.03297	.00372	.01710	.03484
9	S98T002604	.00485	.00036	.01480	.00338
10	S98T002604_D	.00323	-.00001	.02205	.00084
11	S98T002604_S	4.9883	.01052	4.8908	4.7229
12	S98T002604_X	.17247	.03784	.02200	.15331
13	S98T002604_AX	480.94	.83069	465.31	477.01
14	CCV	4.9550	.00823	4.7279	4.8168
15	CCB	.00263	.00021	.00018	.00201
16	S98T002605	.00901	.00071	.01085	.01242
17	S98T002605_D	.01227	.00149	.03074	.01166
18	S98T002606	.02523	.00294	.03864	.02885
19	S98T002606_D	.03558	.00477	.05018	.04385
20	ICSA	.00180	.00772	-.00442	-.00262
21	ICSAB	.45964	.00764	.91682	-.00284
22	CCV	4.9643	.00824	4.7417	4.8216
23	CCB	.00116	.00000	-.00030	.00152

JK 12.21.98

LABCORE Data Entry Template for Worklist# 27672

Analyst: JK Sefto Instrument: ICP01A 12-30-98 Book# 77848

Method: LA-505-151/161 Rev/Mod C-3

Worklist Comment: 12-30-98 ICP BY-105 (SOLID ACID DIGEST)

Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysis

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTD TJA		@ICP-A01	SOLID		
7	PREPBKLTJA		@ICP-A01	SOLID		
8	SAMPLE	S98T002713 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01, MG-A-01, MN-A-01, MO-A-01, NA-A-01, ND-A-01, NI-A-01, P-A-01, PB-A-01, S-A-01, SE-A-01, SI-A-01, SM-A-01, SR-A-01, TI-A-01, TL-A-01, U-A-01, V-A-01, ZN-A-01, ZR-A-01						
9	DUP	S98T002713 0 A	@ICP-A01	SOLID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002846 0 A	@ICP-A01	SOLID		
13	SAMPLE	S98T002846 0 A	@ICP-A01	SOLID	98000437	BY-105 (2)
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01, MG-A-01, MN-A-01, MO-A-01, NA-A-01, ND-A-01, NI-A-01, P-A-01, PB-A-01, S-A-01, SE-A-01, SI-A-01, SM-A-01, SR-A-01, TI-A-01, TL-A-01, U-A-01, V-A-01, ZN-A-01, ZR-A-01						

Data Entry Comments:

uploaded 1-12-99
JK Wauell

Validated by:
Saul M. Pany
1/13/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27672

S Type	Sample#	R A	Test	Matrix	Group#	Project
			SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01			
14 DUP	S98T002846	0 A	@ICP-A01	SOLID		
15 SPK-PREDIG	S98T002846	0 A	@ICP-A01	SOLID		
16 CCV	5646/2 K Spk 6X (10 ppb) 12-30-98		@ICP-QC	QC		
17 CCB			@ICP-QC	QC		
18 SERDIL	S98T002740	0 A	@ICP-A01	SOLID		
19 SAMPLE	S98T002740	0 A	@ICP-A01	SOLID	98000490	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
20 DUP	S98T002740	0 A	@ICP-A01	SOLID		
21 SPK-PREDIG	S98T002740	0 A	@ICP-A01	SOLID		
22 ICBSA	Sample 6X (10 ppb) 12-3-98		@ICP-QC	QC		
23 ICBSAB			@ICP-QC	QC		
24 CCV			@ICP-QC	QC		
25 CCB			@ICP-QC	QC		

Final page for worklist # 27672

Analyst Signature	Date	Analyst Signature	Date
Preps 1d JTG, 3044, 12/1	5987001844-1, 2-8-2-8, 5	5987002740-1, 2-8-2-8, 5	5987002740-1, 2-8-2-8, 5
Preps 1d JTG, 3044, 1	5987001844-2, 2-8-2-8, 5	5987002740-2, 2-8-2-8, 5	5987002740-2, 2-8-2-8, 5
5987001713, 2-8, 12/5	5987001844-3, 2-8-2-8, 5	5987002740-3, 2-8-2-8, 5	5987002740-3, 2-8-2-8, 5
5987001713, 2-8, 5	5987001844-4, 2-8-1-9, 50	5987002740-4, 2-8-1-9, 50	5987002740-4, 2-8-1-9, 50
	5987001844-5, 2-8-1-9, 50		

Data Entry Comments:

S98T002846

HNF-1663 REV. 0

$$Cr = 4.2301 \frac{\mu g}{ml} \times \frac{1}{4.904 \frac{g}{L} \times \frac{1L}{1000 ml}} = 8.63 \pm 2 \frac{\mu g}{g}$$

Pre-dig Spk: $Cr = \frac{8.9276 - 4.2301}{5} \times 100 = 94.0\%$

Post Spk: $Al = \frac{\left(\frac{560.60}{50}\right) - \left(\frac{76.715}{50}\right)}{10} \times 100 = 96.8\%$

$$Na = \frac{\left(\frac{1597.7}{50}\right) - \left(\frac{1127.0}{50}\right)}{10} \times 100 = 94.1\%$$

$$P = \frac{\left(\frac{520.06}{50}\right) - \left(\frac{23.957}{50}\right)}{10} \times 100 = 99.2\%$$

$$Si = \frac{\left(\frac{498.35}{50}\right) - \left(\frac{7.0307}{50}\right)}{10} \times 100 = 98.3\%$$

S98T002740

Post Spk:

$$Al = \frac{\left(\frac{257.47}{50}\right) - \left(\frac{263.25}{50}\right)}{10} \times 100 = 98.8\%$$

$$Na = \frac{\left(\frac{1597.7}{50}\right) - \left(\frac{1127.0}{50}\right)}{10} \times 100 = 94.1\%$$

$$P = \frac{\left(\frac{536.65}{50}\right) - \left(\frac{34.056}{50}\right)}{10} \times 100 = 100.5\%$$

$$Si = \frac{\left(\frac{502.51}{50}\right) - \left(\frac{5.000}{50}\right)}{10} \times 100 = 99.5\%$$

1958

Analysis Report

Summary

Tue 01-12-99 12:41:53 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990112A	ICP2	01/12/99	09:33	DKS	Q	CONC
2	ICB	990112A	ICP2	01/12/99	09:37	DKS	Q	CONC
3	LLS	990112A	ICP2	01/12/99	09:39	DKS	Q	CONC
4	ICSA	990112A	ICP2	01/12/99	09:42	DKS	Q	CONC
5	ICSAB	990112A	ICP2	01/12/99	09:45	DKS	Q	CONC
6	PREPSTDITJA	990112A	ICP2	01/12/99	09:48	DKS	Q	CONC
7	PREPBLKTJA	990112A	ICP2	01/12/99	09:52	DKS	Q	CONC
8	PREPBLKTJA	990112A	ICP2	01/12/99	09:57	DKS	Q	CONC
9	S98T002713	990112A	ICP2	01/12/99	10:00	DKS	S	CONC
10	S98T002713_D	990112A	ICP2	01/12/99	10:04	DKS	S	CONC
11	CCV	990112A	ICP2	01/12/99	10:08	DKS	Q	CONC
12	CCB	990112A	ICP2	01/12/99	10:12	DKS	Q	CONC
13	S98T002846_L	990112A	ICP2	01/12/99	10:15	DKS	S	CONC
14	S98T002846	990112A	ICP2	01/12/99	10:18	DKS	S	CONC
15	S98T002846_D	990112A	ICP2	01/12/99	10:22	DKS	S	CONC
16	S98T002846_S	990112A	ICP2	01/12/99	10:26	DKS	S	CONC
17	S98T002846_X	990112A	ICP2	01/12/99	10:30	DKS	S	CONC
18	S98T002846_AX	990112A	ICP2	01/12/99	10:34	DKS	S	CONC
19	CCV	990112A	ICP2	01/12/99	11:48	DKS	Q	CONC
20	CCB	990112A	ICP2	01/12/99	11:51	DKS	Q	CONC
21	S98T002740_L	990112A	ICP2	01/12/99	11:55	DKS	S	CONC
22	S98T002740	990112A	ICP2	01/12/99	11:58	DKS	S	CONC
23	S98T002740_D	990112A	ICP2	01/12/99	12:02	DKS	S	CONC
24	S98T002740_S	990112A	ICP2	01/12/99	12:05	DKS	S	CONC
25	S98T002740_X	990112A	ICP2	01/12/99	12:09	DKS	S	CONC
26	S98T002740_AX	990112A	ICP2	01/12/99	12:12	DKS	S	CONC
27	ICSA	990112A	ICP2	01/12/99	12:29	DKS	Q	CONC
28	ICSAB	990112A	ICP2	01/12/99	12:32	DKS	Q	CONC
29	CCV	990112A	ICP2	01/12/99	12:35	DKS	Q	CONC
30	CCB	990112A	ICP2	01/12/99	12:38	DKS	Q	CONC

JK Lab

01-12-99

Work/154# 27672

B9-105

598T002713

598T002846

598T002740

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1959 TO 1965.

Analysis Report

Averages

Tue 01-12-99 12:41:53 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9502	4.8460	5.0426	5.2150	4.7567	5.0452
2	ICB	-.00236	-.00790	.01915	.00000	-.00016	-.00018
3	LLS	Q.01491	.09097	.21212	.09832	.09886	.01032
4	ICSA	.00041	242.47	-.00228	-.00616	.00011	.00010
5	ICSAB	.96137	243.47	.01157	-.00694	.45817	.48118
6	PREPSTDITJA	.86265	4.8240	4.6751	5.3685	4.6475	5.1099
7	PREPBLKTJA	Q.00202	Q.14609	Q.01895	Q.73106	Q.00072	Q.00022
8	PREPBLKTJA	-.00134	Q.14277	.02541	Q.72670	.00077	-.00018
9	S98T002713	.09090	102.01	-.02231	.69612	.25214	.00020
10	S98T002713_D	.07752	108.28	.06846	.86664	.29933	.00127
11	CCV	5.0382	4.9115	5.1300	5.3117	4.8352	5.1638
12	CCB	.00029	-.00134	.01351	-.00241	-.00004	-.00004
13	S98T002846_L	.00355	77.955	.44854	.23594	.12358	-.00170
14	S98T002846	.07888	75.011	.02741	.43171	.12320	-.00012
15	S98T002846_D	.07776	79.264	-.02451	.51901	.13445	.00055
16	S98T002846_S	1.0408	73.480	5.2328	5.7221	4.8859	5.1993
17	S98T002846_X	.01325	76.715	.85210	-.01799	.12176	-.00735
18	S98T002846_AX	403.34	560.60	506.46	492.09	491.10	495.85
19	CCV	4.9772	4.8589	5.0516	5.2479	4.7733	5.0917
20	CCB	-.00082	-.00548	.01154	-.00435	-.00002	-.00004
21	S98T002740_L	.06344	260.70	.41232	.50971	.01104	.00000
22	S98T002740	.08009	253.37	.07353	.65964	.01259	-.00043
23	S98T002740_D	.08920	293.95	.17474	.31334	.01546	.00090
24	S98T002740_S	1.0354	263.91	5.2129	5.6788	4.6986	5.1168
25	S98T002740_X	-.00496	263.25	.98174	.50991	.00556	-.00447
26	S98T002740_AX	316.57	757.47	512.36	500.04	501.87	505.94
27	ICSA	.00880	241.98	.02851	-.01991	.00021	.00019
28	ICSAB	.96213	244.86	.01749	-.01326	.46276	.48367
29	CCV	4.9527	4.8543	5.0408	5.2407	4.7853	5.0837
30	CCB	.00179	-.00267	.00702	-.00532	-.00016	.00000

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9904	5.1734	5.0897	4.9662	5.0313	4.9818
2	ICB	-.02481	-.00477	-.00121	-.01252	-.00106	-.00046
3	LLS	.21694	.21297	.00978	.19182	.03957	.01959
4	ICSA	.00310	258.08	.00532	.00043	.00170	-.00201
5	ICSAB	.02156	260.16	.96104	-.00933	.46765	.47538
6	PREPSTDITJA	4.5865	5.2514	4.6552	4.8598	4.6977	4.7113
7	PREPBLKTJA	Q.01335	Q.15462	Q.00125	Q.00897	Q.00105	Q.00303
8	PREPBLKTJA	.00158	Q.15660	-.00028	-.00172	-.00069	.00416
9	S98T002713	1.4183	69.920	.05775	-.00613	.03070	14.863
10	S98T002713_D	1.6835	88.661	.06533	-.00417	.04922	15.752
11	CCV	5.1350	5.2536	5.1935	5.0552	5.1377	5.0819
12	CCB	-.01228	-.00047	.00009	-.00256	.00162	-.00040
13	S98T002846_L	1.1135	10.850	-.01442	-.45311	.03064	4.3787
14	S98T002846	.84789	10.708	.01623	-.03788	.01297	4.2301
15	S98T002846_D	.74662	16.795	.01471	-.05718	.02430	4.3245
16	S98T002846_S	5.7952	16.321	5.2089	5.0421	5.1101	8.9275
17	S98T002846_X	1.7408	10.809	.04468	-.43293	.05474	4.4280
18	S98T002846_AX	508.89	521.86	502.79	498.95	500.73	508.73
19	CCV	5.0610	5.1770	5.1314	5.0230	5.0656	5.0311

Analysis Report

Averages

Tue 01-12-99 12:41:53 PM

page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
20	CCB	.00901	-.00417	-.00077	-.00754	-.00025	-.00023
21	S98T002740_L	.36198	.63464	.03684	-.03975	.02687	1.8694
22	S98T002740	-.07539	.72117	.02167	-.03062	.02738	1.7835
23	S98T002740_D	.00560	.87317	.02976	.00584	.02067	2.1262
24	S98T002740_S	4.9114	6.0359	5.1123	4.9849	5.0323	6.9029
25	S98T002740_X	-.71890	.40872	.01009	-.49191	.04073	1.9058
26	S98T002740_AX	515.07	515.60	509.73	511.29	509.30	513.45
27	ICSA	-.00738	255.48	.00595	-.00170	-.00004	-.00044
28	ICSAB	.00027	257.81	.96036	-.00852	.46431	.47239
29	CCV	5.0502	5.1533	5.1003	5.0184	5.0467	5.0058
30	CCB	.00922	-.00536	-.00047	-.01020	.00174	.00020

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7074	.00695	4.7883	5.0683	4.9704	4.8874
2	ICB	.00018	.00061	-.00230	-.17673	-.00159	.00051
3	LLS	.01629	.00025	.09606	Q.36006	.10148	.02124
4	ICSA	.00424	-.00867	92.928	.04840	-.00502	.00269
5	ICSAB	.46519	-.00959	93.203	.04900	-.00506	.98970
6	PREPSTD TJA	4.4293	.00606	4.5578	4.8104	4.8547	4.8828
7	PREPBLKTJA	Q.00200	.00696	Q.03843	Q.07127	Q.00153	Q.00092
8	PREPBLKTJA	.00302	-.00112	.03300	.09436	-.00046	-.00082
9	S98T002713	-.07264	.07944	33.911	6.4687	.00998	.01191
10	S98T002713_D	-.08137	.09779	35.239	7.5148	.01411	.00863
11	CCV	4.7866	.00688	4.8579	5.0565	5.0727	4.9797
12	CCB	-.00027	.00129	-.00156	-.07868	-.00038	.00102
13	S98T002846_L	-.04255	.02367	10.254	5.3129	-.05071	.02257
14	S98T002846	-.01342	.01448	9.8258	7.8001	.00677	.01436
15	S98T002846_D	-.01690	.02029	10.165	6.3522	.00034	.01796
16	S98T002846_S	4.6155	.02049	14.206	10.841	5.0587	4.9624
17	S98T002846_X	.00079	.08589	10.155	18.825	-.00744	.05604
18	S98T002846_AX	459.15	.65201	489.58	518.93	495.87	482.38
19	CCV	4.7239	.00686	4.8106	5.0109	5.0293	4.9755
20	CCB	.00006	.00099	-.00146	-.00003	-.00090	.00082
21	S98T002740_L	.04846	.01622	3.0573	7.3105	-.00041	.02314
22	S98T002740	.03679	.00246	2.9618	3.7248	-.00088	.00304
23	S98T002740_D	.04328	.00424	2.4639	5.0000	.00497	.00355
24	S98T002740_S	4.6331	.00979	6.9282	7.4608	5.0114	4.9092
25	S98T002740_X	.08046	.05628	3.0166	4.1018	-.04630	.05663
26	S98T002740_AX	471.23	.62863	490.03	490.30	509.13	503.11
27	ICSA	.00438	-.00683	92.343	.32825	-.00412	.00258
28	ICSAB	.46930	-.00572	93.106	-.12430	-.00520	1.0067
29	CCV	4.7290	.00672	4.7799	4.7830	5.0349	5.0100
30	CCB	.00004	.00068	-.00112	-.04975	-.00123	.00082

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1392	4.9275	5.2697	4.8898	4.8395	4.9315
2	ICB	-.01083	.00015	.00367	-.00638	-.00044	.00000
3	LLS	.19933	.02042	.10098	.20123	.20187	.03814
4	ICSA	247.53	-.00427	-.01791	191.93	-.00134	-.01000
5	ICSAB	247.53	.45197	-.01811	191.47	-.00129	.93681

Analysis Report

Averages

Tue 01-12-99 12:41:53 PM

page 4

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
6	PREPSTDITJA	4.9457	4.6098	4.7814	6.4577	4.7981	4.5881
7	PREPBLKTJA	~~2.08056~~	~~2.08084~~	~~2.00102~~	~~21.1564~~	~~2.00096~~	~~2.00598~~
8	PREPBLKTJA	.07197	.00075	.00189	21.1539	.00036	.00176
9	S98T002713	4.8424	.39779	.03336	1164.5	.05657	8.3499
10	S98T002713_D	4.6652	.40678	.03866	1066.2	.08353	9.6051
11	CCV	5.1726	5.0093	5.3769	4.9195	4.9461	5.0726
12	CCB	-.00206	.00049	.00195	-.00498	-.00064	.00472
13	S98T002846_L	.48267	.20099	.02895	1153.3	.00731	4.1308
14	S98T002846	.61005	.18541	.04401	1098.7	.05637	4.0061
15	S98T002846_D	.70230	.19036	.04767	1094.0	.05054	4.1727
16	S98T002846_S	5.8274	5.0929	5.1756	1126.1	5.0195	8.8527
17	S98T002846_X	.72732	.22215	-.00983	1127.0	.17953	4.1269
18	S98T002846_AX	509.28	501.70	501.60	1597.7	483.51	507.89
19	CCV	5.1228	4.9435	5.3218	4.9100	4.9037	5.0236
20	CCB	-.00403	.00034	.00246	-.00710	.00077	-.00227
21	S98T002740_L	.25633	.11686	-.02107	1088.7	.02171	.37067
22	S98T002740	.15764	.10791	.02577	1051.1	.01533	.33800
23	S98T002740_D	.21490	.12627	.02239	1151.2	.01411	.37910
24	S98T002740_S	5.2819	4.9339	5.0985	1125.6	4.9153	5.3174
25	S98T002740_X	-.16431	.13658	-.00730	1095.1	.01473	.44769
26	S98T002740_AX	519.83	508.76	511.53	1601.7	495.05	513.08
27	ICSA	245.86	-.00449	-.01598	191.39	.00112	.00119
28	ICSAB	247.73	.45114	-.01845	192.76	-.00426	.93779
29	CCV	5.0943	4.9233	5.3245	4.9304	4.9013	5.0098
30	CCB	-.00919	.00031	.00280	-.00497	.00106	-.00400

1-12-99

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.2262	5.0355	5.0920	5.3002	4.8284	5.3548
2	ICB	.00604	-.00098	.00435	-.00085	-.00110	.00539
3	LLS	.38534	.20634	.19331	.09930	.19448	.11421
4	ICSA	.02421	.02074	-.03597	.01030	.03799	.00208
5	ICSAB	.02513	1.0088	-.02134	.00101	.06304	.00288
6	PREPSTDITJA	4.6251	4.5257	4.4696	4.5859	4.6263	4.2864
7	PREPBLKTJA	~~2.03538~~	~~2.04032~~	~~2.13101~~	~~2.00325~~	~~2.02681~~	~~2.1714~~
8	PREPBLKTJA	.03668	.00338	.015037	.00327	.01319	2.1295
9	S98T002713	97.178	1.8467	44.214	.01777	.13712	8.6465
10	S98T002713_D	139.14	1.5534	46.036	.00731	.05404	9.2061
11	CCV	5.3257	5.1381	5.1733	5.3964	4.9826	5.4540
12	CCB	.00452	.00429	.00010	.00595	.02869	.00697
13	S98T002846_L	22.596	.27930	5.7016	.01941	.44429	5.1534
14	S98T002846	21.769	.46462	5.4647	.00973	.16442	4.0255
15	S98T002846_D	34.601	.44855	5.7657	-.00267	.21791	5.6887
16	S98T002846_S	21.174	5.5152	10.260	5.2036	5.2419	10.494
17	S98T002846_X	23.957	.85759	5.8688	.04984	1.6278	7.0307
18	S98T002846_AX	520.06	511.36	496.49	503.43	490.48	498.35
19	CCV	5.2150	5.0706	5.0880	5.3428	4.8837	5.3623
20	CCB	.00474	.00863	.00882	.00364	.04508	.00051
21	S98T002740_L	33.048	.42278	7.8157	-.00459	1.3227	3.8052
22	S98T002740	32.054	.23617	7.8720	.00631	.46749	2.5450
23	S98T002740_D	30.100	.26175	9.5436	.01915	.45378	2.6655
24	S98T002740_S	45.275	5.1802	13.834	5.0852	5.4060	8.8513
25	S98T002740_X	34.056	.42553	8.0334	.01782	2.8010	5.0000

1-12-99

Analysis Report

Averages

Tue 01-12-99 12:41:53 PM

page 5

#	Sample Name	P	Pb	S	Sb	Se	Si
26	S98T002740_AX	536.65	517.58	504.97	514.19	499.79	502.51
27	ICSA	.05547	.02155	-.02425	.00110	.10459	-.00108
28	ICSAB	.04760	1.0103	-.02499	-.00849	.07033	-.00372
29	CCV	5.2866	5.0788	5.1045	5.3434	4.9051	5.3463
30	CCB	.02068	.00355	-.01467	-.00085	.07281	.00103

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8225	4.8415	.19725	5.2668	4.6927	9.7034
2	ICB	.00182	.00000	-.01371	-.00028	-.01248	.01398
3	LLS	.19826	.02030	-.00493	.01979	.40158	.48715
4	ICSA	-.01532	.00180	-.00524	.00095	.04077	-.14689
5	ICSAB	-.01572	.00172	-.00400	.00095	-.01181	-.14432
6	PREPSTDITJA	4.7851	4.8016	.18104	4.8298	4.4547	9.3354
7	PREPBLKTJA	0.01203	0.00030	0.00537	0.00021	0.00330	0.04048
8	PREPBLKTJA	-.01836	.00025	-.00074	.00114	-.01774	-.04548
9	S98T002713	.00968	4.8908	.25248	.79624	.06257	18.426
10	S98T002713_D	.02899	6.3037	.31160	.58095	-.01487	21.784
11	CCV	4.9074	4.9324	.19432	5.3646	4.8785	9.8705
12	CCB	.01241	.00016	-.00789	-.00003	.00216	.05133
13	S98T002846_L	.14095	5.5951	-.09664	.05108	-.22450	12.172
14	S98T002846	-.00104	5.3822	.14402	.05804	-.08683	11.177
15	S98T002846_D	-.00196	5.5430	.15576	.07059	-.06775	11.453
16	S98T002846_S	4.9013	10.029	.34958	5.1208	4.8543	20.081
17	S98T002846_X	.61359	5.4666	-.20083	.09649	.17932	13.879
18	S98T002846_AX	478.32	494.45	19.545	491.43	489.24	943.36
19	CCV	4.8680	4.8876	.19934	5.2919	4.7818	9.7808
20	CCB	.00588	.00003	-.00729	.00065	.02212	.03411
21	S98T002740_L	.08896	.14740	-.06147	.06233	.23804	1.5410
22	S98T002740	.00559	.14274	-.00446	.04781	-.07424	.90153
23	S98T002740_D	.01647	.17074	.01479	.07540	-.02860	1.0706
24	S98T002740_S	4.8752	5.0809	.18283	5.0814	4.7630	10.427
25	S98T002740_X	.35944	.15371	-.49224	.04320	1.1705	2.8117
26	S98T002740_AX	491.78	501.12	20.138	503.65	499.04	962.26
27	ICSA	-.01667	.00184	.00371	.00143	.00462	-.12669
28	ICSAB	-.02058	.00180	-.00448	.00048	.00437	-.13596
29	CCV	4.8676	4.8838	.19334	5.2808	4.7801	9.8193
30	CCB	.00316	.00004	-.00932	.00018	.00996	.01882

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0057	.00833	4.8285	5.0638
2	ICB	.00073	.00022	-.00053	.00164
3	LLS	.10064	.00026	.01925	.01977
4	ICSA	.00212	.00774	-.00374	-.00190
5	ICSAB	.47460	.00781	.94953	-.00182
6	PREPSTDITJA	4.6756	.00684	4.2307	4.6452
7	PREPBLKTJA	0.00194	0.00076	0.00727	0.00173
8	PREPBLKTJA	-.00267	-.00084	.00640	.00075
9	S98T002713	.11955	.02155	.14872	.04627
10	S98T002713_D	.25440	.02443	.16729	.05631
11	CCV	5.1009	.00841	4.9227	5.1644

Analysis Report

Averages

Tue 01-12-99 12:41:53 PM

page 6

#	Sample Name	V	Y	Zn	Zr
12	CCB	.00357	.00076	.00003	.00246
13	S98T002846_L	.05816	.01245	.05959	.04456
14	S98T002846	.03289	.00621	.05919	.01791
15	S98T002846_D	.04129	.00610	.05365	.01504
16	S98T002846_S	5.0115	.01418	4.9369	4.9005
17	S98T002846_X	.21062	.04894	.03908	.13067
18	S98T002846_AX	495.20	.81745	475.16	494.22
19	CCV	5.0464	.00814	4.8605	5.1114
20	CCB	.00159	.00045	-.00025	.00147
21	S98T002740_L	.11087	.00718	.01025	.05134
22	S98T002740	.07440	.00235	.01699	.02356
23	S98T002740_D	.06592	.00307	.01780	.02521
24	S98T002740_S	5.0057	.01045	4.7978	4.4880
25	S98T002740_X	.23300	.03014	-.00692	.14642
26	S98T002740_AX	504.97	.81725	480.27	506.31
27	ICSA	.00160	.00765	-.00364	-.00315
28	ICSAB	.47337	.00742	.94539	-.00271
29	CCV	5.0292	.00812	4.8217	5.1044
30	CCB	.00126	.00022	-.00011	.00163

JK Lgd

01-12-99

LABCORE Data Entry Template for Worklist# 27674

Analyst: JK Sato Instrument: ICP01 2-12-36-98Book# 77B30Method: LA-505-151/161 Rev/Mod C-3JK 12-30-98

Worklist Comment: ICP BY-105 (SOLID ACID DIGEST)

Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysis

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTDITJA		@ICP-A01	SOLID		
7	PREPBLKTJA		@ICP-A01	SOLID		
8	SERDIL	S98T003394 0 A	@ICP-A01	SOLID		
9	SAMPLE	S98T003394 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
10	DUP	S98T003394 0 A	@ICP-A01	SOLID		
11	SPK ₇ PREDIG	S98T003394 0 A	@ICP-A01	SOLID		
12	CCV	<u>Sample X</u> <u>SpkC 4X (10ppm) JK 12-30-98</u>	@ICP-QC	QC		
13	CCB		@ICP-QC	QC		
14	SERDIL	S98T003399 0 A	@ICP-A01	SOLID		
15	SAMPLE	S98T003399 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 ,						

Data Entry Comments:

uploaded 1-4-99JK Wmell

Validated by:

Sal M. Pang01/06/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27674

S Type	Sample#	R A	Test	Matrix	Group#	Project
			BE-A-01 ,	BI-A-01 ,	CA-A-01 ,	CD-A-01 , CE-A-01 , CO-A-01 ,
			CR-A-01 ,	CU-A-01 ,	FE-A-01 ,	K-A-01 , LA-A-01 , LI-A-01 ,
			MG-A-01 ,	MN-A-01 ,	MO-A-01 ,	NA-A-01 , ND-A-01 , NI-A-01 ,
			P-A-01 ,	PB-A-01 ,	S-A-01 ,	SB-A-01 , SE-A-01 , SI-A-01 ,
			SM-A-01 ,	SR-A-01 ,	TI-A-01 ,	TL-A-01 , U-A-01 , V-A-01 ,
			ZN-A-01 ,	ZR-A-01		
16 DUP	S98T003399	0 A	@ICP-A01	SOLID		
17 SPK-PREDIG	S98T003399	0 A	@ICP-A01	SOLID		
18 CCV	S98T003399		@ICP-QC	QC		
19 CCB			@ICP-QC	QC		
20 SERDIL	S98T003389	0 A	@ICP-A01	SOLID		
21 SAMPLE	S98T003389	0 A	@ICP-A01	SOLID	98000490	BY-105 (2)
	Analytes Requested:		AG-A-01 ,	AL-A-01 ,	AS-A-01 ,	B-A-01 , BA-A-01 ,
			BE-A-01 ,	BI-A-01 ,	CA-A-01 ,	CD-A-01 , CE-A-01 , CO-A-01 ,
			CR-A-01 ,	CU-A-01 ,	FE-A-01 ,	K-A-01 , LA-A-01 , LI-A-01 ,
			MG-A-01 ,	MN-A-01 ,	MO-A-01 ,	NA-A-01 , ND-A-01 , NI-A-01 ,
			P-A-01 ,	PB-A-01 ,	S-A-01 ,	SB-A-01 , SE-A-01 , SI-A-01 ,
			SM-A-01 ,	SR-A-01 ,	TI-A-01 ,	TL-A-01 , U-A-01 , V-A-01 ,
			ZN-A-01 ,	ZR-A-01		
22 DUP	S98T003389	0 A	@ICP-A01	SOLID		
23 SPK-PREDIG	S98T003389	0 A	@ICP-A01	SOLID		
24 ICSA	S98T003389		@ICP-QC	QC		
25 ICSAB			@ICP-QC	QC		
26 CCV			@ICP-QC	QC		
27 CCB			@ICP-QC	QC		

Final page for worklist # 27674

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1966

LABCORE Data Entry Template for Worklist# 27674

S Type	Sample#	R A Test	Matrix	Group#	Project
Prep Std TQ	Direct	DF 1			
Prep Blk TQ	Direct	1			
5987003394-L	2-8-2-8	DF 25			
5987003394	2-8	5			
5987003394-D	2-8	5			
5987003394-S	2-8	5			
5987003394-X	2-8-1-9	50			
5987003394-QX	2-8-1-9	50			
5987003399-L	4-4-2-8	DF 10			
5987003399	4-4	2			
5987003399-D	4-4	2			
5987003399-S	4-4	2			
5987003399-K	4-4-1-9	20			
5987003399-QX	4-4-1-9	20			
5987003387-L	2-8-2-8	DF 25			
5987003387	2-8	5			
5987003387-D	2-8	5			
5987003387-S	2-8	50.5			12-31-98
5987003387-X	2-8-1-9	50			
5987003387-QX	2-8-1-9	50			

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

$$Cr = 0.62869 \frac{\mu g}{ml} \times \frac{1}{5.0889 \frac{g}{L} \times \frac{1L}{1000 ml}} = 1.24 \times 10^{-2} \frac{\mu g}{g}$$

Pre-dig Spk:

$$Cr = \frac{5.7085 - 0.62869}{5} \times 100 = 101.6\%$$

Post Spk:

$$Al = \frac{\left(\frac{533.07}{50}\right) - \left(\frac{41.485}{50}\right)}{10} \times 100 = 98.3\%$$

$$Fe = \frac{\left(\frac{496.56}{50}\right) - \left(\frac{10.447}{50}\right)}{10} \times 100 = 97.2\%$$

$$Na = \frac{\left(\frac{1783.7}{50}\right) - \left(\frac{1296.6}{50}\right)}{10} \times 100 = 97.4\%$$

S98T003389

$$Cr = 1.3826 \frac{\mu g}{ml} \times \frac{1}{4.8789 \frac{g}{L} \times \frac{1L}{1000 ml}} = 2.83 \times 10^{-2} \frac{\mu g}{g}$$

Pre-dig Spk:

$$Cr = \frac{6.4601 - 1.3826}{5} \times 100 = 101.6\%$$

Post Spk:

$$Al = \frac{\left(\frac{562.27}{50}\right) - \left(\frac{78.522}{50}\right)}{10} \times 100 = 96.7\%$$

$$Na = \frac{\left(\frac{1640.5}{50}\right) - \left(\frac{1161.2}{50}\right)}{10} \times 100 = 95.9\%$$

$$C_2 = 2.3868 \frac{\mu\text{g}}{\text{mL}} \times \frac{1}{4.986 \frac{\text{g}}{\text{L}} \times \frac{1\text{L}}{1000\text{mL}}} = 4.79 \pm 2 \frac{\mu\text{g}}{\text{g}}$$

Pre-dig Spk:

$$C_2 = \frac{7.0789 - 2.3868}{5} \times 100 = 93.8\%$$

Post Spk:

$$Al = \frac{\left(\frac{304.59}{20}\right) - \left(\frac{109.21}{20}\right)}{10} \times 100 = 97.7\%$$

$$Cu = \frac{\left(\frac{187.56}{20}\right) - 0}{10} \times 100 = 93.8\%$$

$$Fe = \frac{\left(\frac{331.44}{20}\right) - \left(\frac{143.76}{20}\right)}{10} \times 100 = 93.8\%$$

$$Mn = \frac{\left(\frac{823.64}{20}\right) - \left(\frac{625.10}{20}\right)}{10} \times 100 = 99.3\%$$

$$Ni = \frac{\left(\frac{259.18}{20}\right) - \left(\frac{61.442}{20}\right)}{10} \times 100 = 98.9\%$$

$$P = \frac{\left(\frac{295.03}{20}\right) - \left(\frac{98.380}{20}\right)}{10} \times 100 = 98.3\%$$

80
01/06/98

$$V = \frac{\left(\frac{198.13}{20}\right) - (0)}{10} \times 100 = 99.07\%$$

$$Zr = \frac{\left(\frac{199.05}{20}\right) - (0)}{10} \times 100 = 99.5\%$$

Analysis Report

Summary

Thu 12-31-98 02:16:05 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	981231B	ICP2	12/31/98	10:12	DKS	Q	CONC
2	ICB	981231B	ICP2	12/31/98	10:15	DKS	Q	CONC
3	LLS	981231B	ICP2	12/31/98	10:18	DKS	Q	CONC
4	ICSA	981231B	ICP2	12/31/98	10:21	DKS	Q	CONC
5	ICSAB	981231B	ICP2	12/31/98	10:24	DKS	Q	CONC
6	PREPSTD TJ A	981231B	ICP2	12/31/98	10:27	DKS	Q	CONC
7	PREPBLK TJ A	981231B	ICP2	12/31/98	10:30	DKS	Q	CONC
8	S98T003394_L	981231B	ICP2	12/31/98	10:34	DKS	S	CONC
9	S98T003394	981231B	ICP2	12/31/98	10:37	DKS	S	CONC
10	S98T003394_D	981231B	ICP2	12/31/98	10:41	DKS	S	CONC
11	S98T003394_S	981231B	ICP2	12/31/98	10:44	DKS	S	CONC
12	S98T003394_X	981231B	ICP2	12/31/98	10:49	DKS	S	CONC
13	S98T003394_AX	981231B	ICP2	12/31/98	10:52	DKS	S	CONC
14	CCV	981231B	ICP2	12/31/98	12:06	DKS	Q	CONC
15	CCB	981231B	ICP2	12/31/98	12:10	DKS	Q	CONC
16	S98T003399_L	981231B	ICP2	12/31/98	12:13	DKS	S	CONC
17	S98T003399_L	981231B	ICP2	12/31/98	12:19	DKS	S	CONC
18	S98T003399	981231B	ICP2	12/31/98	12:23	DKS	S	CONC
19	S98T003399_D	981231B	ICP2	12/31/98	12:26	DKS	S	CONC
20	S98T003399_S	981231B	ICP2	12/31/98	12:29	DKS	S	CONC
21	S98T003399_X	981231B	ICP2	12/31/98	12:35	DKS	S	CONC
22	S98T003399_AX	981231B	ICP2	12/31/98	12:38	DKS	S	CONC
23	CCV	981231B	ICP2	12/31/98	13:04	DKS	Q	CONC
24	CCB	981231B	ICP2	12/31/98	13:12	DKS	Q	CONC
25	S98T003389_L	981231B	ICP2	12/31/98	13:16	DKS	S	CONC
26	S98T003389	981231B	ICP2	12/31/98	13:19	DKS	S	CONC
27	S98T003389_D	981231B	ICP2	12/31/98	13:22	DKS	S	CONC
28	S98T003389_S	981231B	ICP2	12/31/98	13:26	DKS	S	CONC
29	S98T003389_X	981231B	ICP2	12/31/98	13:30	DKS	S	CONC
30	S98T003389_AX	981231B	ICP2	12/31/98	13:34	DKS	S	CONC
31	ICSA	981231B	ICP2	12/31/98	14:03	DKS	Q	CONC
32	ICSAB	981231B	ICP2	12/31/98	14:06	DKS	Q	CONC
33	CCV	981231B	ICP2	12/31/98	14:09	DKS	Q	CONC
34	CCB	981231B	ICP2	12/31/98	14:12	DKS	Q	CONC

12-31-98
Work/ist # 27674

B 4-105
598T003394
598T003399
598T003389

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1970 TO 1975.

Analysis Report

Averages

Thu 12-31-98 02:16:05 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	5.0013	4.8797	5.0775	5.1793	4.8379	5.1353
2	ICB	.00002	.00019	.00034	.00292	.00008	-.00004
3	LLS	.02049	.09904	.19801	.10132	.09831	.01025
4	ICSA	.00140	243.85	.03916	-.00868	.00025	.00000
5	ICSAB	.95536	243.92	.04451	-.00565	.45977	.47947
6	PREPSTDTJA	.84965	4.7282	4.6144	5.3010	4.6804	5.1495
7	PREPLKTJA	-.00219	Q.13246	-.00397	Q.76058	.00053	-.00009
8	S98T003394_L	.05618	42.130	-.23994	.77111	.00577	-.00234
9	S98T003394	.09114	40.125	.01738	.72928	.00442	-.00070
10	S98T003394_D	.08264	46.299	-.01970	.76228	.00249	-.00023
11	S98T003394_S	1.0408	46.769	5.0714	5.7811	4.7738	5.1554
12	S98T003394_X	.02628	41.685	-.11450	.85627	.00385	.00233
13	S98T003394_AX	462.22	533.07	507.60	504.61	502.97	502.72
14	CCV	4.9805	4.8538	5.0146	5.1053	4.7979	5.0838
15	CCB	.00042	-.00672	.00885	.00779	.00013	-.00009
16	S98T003399_L	.18999	267.95	.24959	1.5757	5.3698	.00224
17	S98T003399_L	.07417	107.28	.04595	.68405	2.1572	.00137
18	S98T003399	.08460	101.00	-.05796	.58337	2.0158	.00190
19	S98T003399_D	.08906	118.48	-.02229	.65277	2.3681	.00205
20	S98T003399_S	.93424	118.95	4.6081	5.1448	6.6470	4.7227
21	S98T003399_X	.05321	109.21	-.01822	.64192	2.1922	-.00061
22	S98T003399_AX	167.11	304.59	202.03	200.85	204.28	200.65
23	CCV	5.0019	4.9182	5.0497	5.1564	4.8890	5.1138
24	CCB	.00146	-.01469	-.00788	-.00340	.00000	-.00032
25	S98T003389_L	.12296	78.984	.21325	.57245	.00685	-.00580
26	S98T003389	.08874	76.339	-.00247	.54084	.00643	.00000
27	S98T003389_D	.08698	79.062	-.00455	.72831	.00595	-.00047
28	S98T003389_S	1.0354	87.648	4.9616	5.7693	4.6207	4.9629
29	S98T003389_X	.06989	78.522	.16072	.51201	.00587	-.01160
30	S98T003389_AX	371.90	562.27	503.37	496.53	496.21	494.13
31	ICSA	.01041	240.65	.03985	-.01346	.00008	-.00004
32	ICSAB	.95992	241.84	.03073	-.00676	.45720	.47285
33	CCV	5.0141	4.8779	5.0534	5.1544	4.8538	5.1031
34	CCB	.00555	-.00608	.00299	.00146	.00000	-.00013

12-31-98

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9922	5.2513	5.1732	5.0220	5.0457	5.0607
2	ICB	.00709	.00387	.00001	.00746	.00001	.00195
3	LLS	.18277	.22310	.01049	.20155	.04066	.02190
4	ICSA	-.02084	257.46	.00653	-.00183	.00023	-.00094
5	ICSAB	-.01339	257.33	.95331	.00554	.46689	.47448
6	PREPSTDTJA	4.4898	4.8765	4.6469	4.8580	4.6668	4.7030
7	PREPLKTJA	.00677	Q.16505	-.00046	-.00612	.00023	.00393
8	S98T003394_L	-.39612	.50966	.01269	-.02245	.01031	.71567
9	S98T003394	.08474	.40549	.02425	.06440	.01561	.62869
10	S98T003394_D	-.06856	.34977	.00493	-.01810	.00069	.70797
11	S98T003394_S	4.9562	5.7359	5.1815	5.0254	5.0877	5.7085
12	S98T003394_X	.22305	.39511	-.01125	-.30058	-.00603	.67231
13	S98T003394_AX	508.64	521.00	509.08	504.20	504.02	512.13
14	CCV	5.0170	5.2288	5.1391	4.9676	5.0132	5.0330
15	CCB	-.02010	.00339	-.00074	.00147	.00158	.00182

Analysis Report Averages Thu 12-31-98 02:16:05 PM page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
16	S98T003399_L	38.361	307.23	.01671	.17660	.25421	6.4163
17	S98T003399_L	15.180	123.39	-.00005	-.11122	.09125	2.5861
18	S98T003399	14.284	115.17	-.00064	-.04108	.08409	2.3868
19	S98T003399_D	16.922	130.81	.00161	-.03096	.09357	2.7338
20	S98T003399_S	20.460	131.74	4.4597	4.4606	4.3036	7.0789
21	S98T003399_X	15.167	126.41	-.01962	-.12232	.04841	2.6401
22	S98T003399_AX	216.87	325.10	200.06	201.76	199.43	203.35
23	CCV	5.0544	5.2161	5.1391	5.0348	5.0336	5.0402
24	CCB	-.00841	-.00584	-.00101	-.01378	-.00115	.00008
25	S98T003389_L	-.28772	.68557	.02051	.07616	.01276	1.4442
26	S98T003389	-.01129	.59071	.00663	-.00465	.00638	1.3826
27	S98T003389_D	-.07897	.51796	.00965	.00725	.01289	1.4581
28	S98T003389_S	4.8923	5.8273	5.0316	4.8453	4.9524	6.4601
29	S98T003389_X	-1.0818	.78265	-.00366	-.03006	.00065	1.4494
30	S98T003389_AX	506.32	509.85	500.53	496.82	497.79	503.33
31	ICSA	.00792	255.29	.00528	-.00394	.00309	-.00154
32	ICSAB	-.01445	258.63	.94820	.00768	.46560	.47230
33	CCV	5.0799	5.2580	5.1568	5.0040	5.0390	5.0533
34	CCB	-.01001	.00850	-.00041	-.00401	-.00180	-.00018

12-31-98

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7180	.00739	4.8486	4.8613	5.0275	4.8803
2	ICB	.00105	.00101	-.00022	-.22741	.00030	.00051
3	LLS	.01804	.00104	.09784	.41922	.10073	.02120
4	ICSA	.00333	.00147	93.279	.00048	-.00426	.00270
5	ICSAB	.46443	.00486	93.318	-.02252	-.00381	.96794
6	PREPSTDITJA	4.4182	.00536	4.5727	4.5269	4.8486	4.7751
7	PREPBLKTJA	.00666	-.00105	.04707	-.29851	-.00154	-.00093
8	S98T003394_L	.03885	.03882	10.701	-1.7374	-.00576	.02074
9	S98T003394	.02044	.00552	10.072	2.3353	.00701	.00987
10	S98T003394_D	.01548	.00515	.54406	2.1204	-.00402	.00727
11	S98T003394_S	4.6812	.00878	5.2584	5.5069	4.9990	4.8632
12	S98T003394_X	.04425	.04438	10.447	-10.082	-.03437	.01553
13	S98T003394_AX	466.06	.67571	496.56	497.49	501.30	490.91
14	CCV	4.6700	.00699	4.8274	4.7725	4.9765	4.8926
15	CCB	.00063	.00190	-.00042	-.12320	-.00041	.00134
16	S98T003399_L	1.3320	.36605	350.50	11.749	.16407	.05110
17	S98T003399_L	-.52915	.12153	140.65	4.8349	.07340	.01835
18	S98T003399	-.50592	.07497	129.07	5.5163	.07832	.01496
19	S98T003399_D	-.59694	.07941	151.10	5.9046	.09670	.01723
20	S98T003399_S	2.8420	.08427	149.09	10.536	4.6054	4.5505
21	S98T003399_X	-.53738	.14867	143.76	3.3208	.06580	.02550
22	S98T003399_AX	187.56	.36405	331.44	202.51	201.17	197.94
23	CCV	4.7426	.00657	4.8389	5.0312	5.0485	5.0263
24	CCB	-.00037	.00074	.00024	-.16007	-.00134	.00041
25	S98T003389_L	.04398	.00525	.87563	.58913	.00041	.00000
26	S98T003389	.02461	.00217	.82556	2.4878	.00229	.00206
27	S98T003389_D	.01993	.00569	.79768	1.3216	-.00073	.00257
28	S98T003389_S	4.4890	.01450	5.5865	7.3381	4.8252	4.6587
29	S98T003389_X	.03073	.06306	.87873	-2.1039	-.01064	.02589
30	S98T003389_AX	462.34	.63071	478.78	495.20	493.72	483.78
31	ICSA	.00338	.00059	92.161	.02079	-.00508	.00270

12-31-98

Analysis Report   Averages   Thu 12-31-98 02:16:05 PM   page 4

#	Sample Name	Cu	Eu	Fe	K	La	Li
32	ICSAB	.45754	-.00070	92.926	.15931	-.00368	.95587
33	CCV	4.7143	.00725	4.8326	4.9444	5.0124	4.9149
34	CCB	.00043	.00067	-.00074	-.18324	-.00057	.00020

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1409	4.9936	5.1176	4.8882	4.9042	5.0248
2	ICB	.00267	.00052	.00017	-.00505	.00331	-.00689
3	LLS	.20536	.02078	.10233	.19564	.20165	.04602
4	ICSA	250.01	-.00471	-.01262	190.71	.00239	-.01156
5	ICSAB	249.92	.44788	-.01794	189.87	.00213	.93167
6	PREPSTDTJA	4.7014	4.6024	4.7467	5.6913	4.7941	4.5831
7	PREPBLKTJA	.01268	.00024	.00167	Q1.1761	.00112	.00099
8	S98T003394_L	-.03515	.09246	-.02928	1293.7	.09465	-.01526
9	S98T003394	.07343	.07648	.00863	1217.2	.01213	.06177
10	S98T003394_D	.02139	.01648	.00862	1209.9	.00920	.04972
11	S98T003394_S	5.2503	4.8992	5.1350	1235.1	4.9220	5.0533
12	S98T003394_X	-.23634	.09483	-.02494	1296.6	.06907	.03708
13	S98T003394_AX	518.28	509.87	503.34	1783.7	489.02	510.65
14	CCV	5.1238	4.9514	5.0833	4.8669	4.8585	4.9887
15	CCB	-.00360	.00066	.00002	-.02308	.00545	-.00024
16	S98T003399_L	16.587	4.5010	-.04099	1540.7	2.0416	149.32
17	S98T003399_L	6.7453	1.8104	-.01461	612.30	.81921	59.809
18	S98T003399	6.2493	1.6703	-.00591	572.36	.69859	54.483
19	S98T003399_D	7.2122	1.9548	-.01023	642.40	.81026	63.037
20	S98T003399_S	11.565	6.2114	4.5096	680.53	5.2279	68.204
21	S98T003399_X	6.7051	1.8580	-.01996	625.10	.79875	61.442
22	S98T003399_AX	213.51	202.47	199.65	823.64	197.22	259.18
23	CCV	5.1860	4.9681	5.1111	4.9924	4.9351	5.0089
24	CCB	-.01129	.00006	-.00135	.06501	-.00121	.00157
25	S98T003389_L	.13862	.04071	-.03414	1175.7	.06557	.20671
26	S98T003389	.09726	.04082	.00865	1120.6	.02017	.29584
27	S98T003389_D	.06095	.03976	.00958	1225.7	.01565	.31330
28	S98T003389_S	5.1330	4.7950	5.0545	1219.3	4.7407	5.1614
29	S98T003389_X	-.07121	.06198	-.03295	1161.2	.06234	.35115
30	S98T003389_AX	513.32	499.95	498.05	1640.5	482.04	500.94
31	ICSA	246.52	-.00457	-.01500	189.20	.00433	-.00949
32	ICSAB	247.74	.44573	-.01556	187.10	.00382	.92838
33	CCV	5.1243	4.9754	5.1209	4.8979	4.8981	5.0388
34	CCB	-.00481	.00015	.00119	-.01406	-.00029	.00025

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.1006	5.0716	4.9612	5.1410	4.9002	5.2070
2	ICB	.01444	.00719	.01191	.00026	.03138	.00554
3	LLS	.41244	.21022	.21038	.18893	.22379	.11690
4	ICSA	.05053	.03483	-.02646	.00439	.07036	-.00224
5	ICSAB	.04905	1.0196	-.02822	.00192	.08102	.00189
6	PREPSTDTJA	4.5632	4.4428	4.4088	4.5244	4.5634	8.1803
7	PREPBLKTJA	.03024	.00244	Q.13070	.00293	.02993	Q2.2303
8	S98T003394_L	2.6635	.27941	1.8187	.28549	.81312	1.8173
9	S98T003394	2.5029	.25638	1.6324	.02723	.14058	1.5299

Analysis Report

Averages

Thu 12-31-98 02:16:05 PM

page 5

#	Sample Name	P	Pb	S	Sb	Se	Si
10	S98T003394_D	2.7847	.07173	1.9159	.05677	.19547	1.8151
11	S98T003394_S	7.5812	5.1160	6.6856	5.1479	5.1144	9.2391
12	S98T003394_X	2.7980	.50326	1.5286	.34121	1.3065	2.2800
13	S98T003394_AX	503.45	514.60	497.25	505.16	496.96	516.33
14	CCV	5.0447	5.0632	4.9325	5.0936	4.8823	5.1787
15	CCB	.01906	.01075	.00963	.00172	.03684	.00877
16	S98T003399_L	241.21	9.0820	20.407	.07350	.40107	37.930
17	S98T003399_L	95.865	3.6442	8.1307	.04262	.26246	15.205
18	S98T003399	89.273	3.3296	7.4941	-.00572	.01378	14.253
19	S98T003399_D	106.12	3.8992	5.0280	-.01152	.00334	11.910
20	S98T003399_S	108.55	8.2377	22.975	4.5549	4.5341	29.071
21	S98T003399_X	98.380	3.5510	8.4575	-.06463	.50110	16.351
22	S98T003399_AX	295.03	204.62	204.62	201.20	198.59	268.42
23	CCV	5.0703	5.0783	4.9539	5.1163	4.8686	5.2048
24	CCB	.01462	-.00254	.00837	.00497	.01707	.00848
25	S98T003389_L	9.1565	.15147	8.7448	.22416	.53588	1.5053
26	S98T003389	8.9254	.12021	8.5036	.02915	.21033	1.3401
27	S98T003389_D	7.7738	.15453	9.4382	.04387	.28142	1.7557
28	S98T003389_S	13.169	5.1068	14.188	5.0012	5.0236	8.1813
29	S98T003389_X	9.5058	.51015	9.3059	.15799	1.6515	1.9313
30	S98T003389_AX	506.76	506.11	501.75	500.63	489.59	492.33
31	ICSA	.04444	.04259	-.01996	.00499	.08959	.00151
32	ICSAB	.05107	1.0194	-.02753	.00825	.09292	.00342
33	CCV	5.1133	5.1045	4.9960	5.1541	4.9246	5.2007
34	CCB	.01110	-.00172	.00721	.00082	.02111	.00621

12-31-98

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8738	4.9118	.17557	5.1091	4.8164	9.6549
2	ICB	.00750	.00012	-.00112	.00068	-.02763	.02717
3	LLS	.20183	.02016	-.00450	.02093	.37003	.49904
4	ICSA	-.01599	.00180	-.00433	.00206	-.05486	-.03197
5	ICSAB	-.01653	.00181	.00327	.00209	-.02020	-.03748
6	PREPSTD TJ A	4.7551	4.8019	.16105	4.8216	4.3819	9.2888
7	PREPBLK TJ A	-.01554	.00004	.00056	.00116	-.02249	-.05603
8	S98T003394_L	.37846	.03738	-.19132	.00496	-.31865	1.3191
9	S98T003394	.05069	.03218	.00386	.00576	-.03342	.26006
10	S98T003394_D	.04418	.03086	-.04081	.00792	-.13410	.23520
11	S98T003394_S	4.8601	4.9774	.16142	5.0861	4.8255	9.5403
12	S98T003394_X	.25309	.03738	-.31312	.02170	-1.5217	1.3961
13	S98T003394_AX	484.10	496.81	17.844	496.58	498.38	938.64
14	CCV	4.8147	4.8660	.16952	5.0637	4.8109	9.5586
15	CCB	.01765	.00021	-.00336	.00090	-.01659	.06672
16	S98T003399_L	.29173	.251.99	6.1422	1.1813	.34859	450.90
17	S98T003399_L	.06523	101.07	2.5704	.47513	-.21397	180.50
18	S98T003399	.05119	100.19	2.2958	.43717	-.08288	169.96
19	S98T003399_D	.05247	118.01	2.6684	.50346	-.15545	201.19
20	S98T003399_S	4.4837	122.35	2.6942	5.0558	4.0709	204.68
21	S98T003399_X	.22500	102.63	2.4334	.47359	-.35394	182.95
22	S98T003399_AX	195.38	304.32	9.8802	199.27	197.41	162.03
23	CCV	4.8936	4.9640	.17973	5.1059	4.7987	9.6879
24	CCB	.00559	.00008	-.00946	-.00026	-.00694	.00849
25	S98T003389_L	-.06718	.05293	-.05267	.04031	-.96781	.09914

12-31-98

Analysis Report

Averages

Thu 12-31-98 02:16:05 PM

page 6

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
26	S98T003389	.00077	.05136	-.00095	.02305	-.24904	.34370
27	S98T003389_D	.02767	.05533	-.00536	.02534	-.13513	.46034
28	S98T003389_S	4.7177	4.8596	.17019	4.9395	4.5132	9.5279
29	S98T003389_X	.44793	.06156	-.28290	.02165	-.31640	2.2817
30	S98T003389_AX	477.48	492.73	17.962	490.95	490.12	929.07
31	ICSA	-.01171	.00187	-.01849	.00247	-.03379	-.01328
32	ICSAB	-.01039	.00194	-.00287	.00277	-.02597	-.03438
33	CCV	4.8486	4.9353	.17648	5.0779	4.8540	9.6196
34	CCB	.00403	.00004	-.00614	-.00002	-.01578	.00714

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0430	.00780	4.8221	4.8998
2	ICB	.00213	.00038	.00089	.00135
3	LLS	.10191	.00064	.02046	.02072
4	ICSA	.00186	.00753	-.00376	-.00166
5	ICSAB	.46968	.00738	.94079	-.00299
6	PREPSTD TJ A	4.6688	.00569	4.1776	4.6181
7	PREPBLK TJ A	-.00300	-.00107	Q.01129	.00001
8	S98T003394_L	.07764	.01903	.03095	.06949
9	S98T003394	.01462	.00184	.03588	.01373
10	S98T003394_D	.01498	.00068	.02804	.01584
11	S98T003394_S	4.9746	.00691	4.8847	4.8802
12	S98T003394_X	.10513	.02291	.09174	.08681
13	S98T003394_AX	498.20	.77661	478.52	495.24
14	CCV	5.0041	.00784	4.8060	4.8611
15	CCB	.00406	.00091	.00086	.00311
16	S98T003399_L	.20320	.17002	1.6000	.13619
17	S98T003399_L	.08063	.06670	.63923	.04923
18	S98T003399	.06214	.06115	.59061	.04588
19	S98T003399_D	.06996	.07138	.68064	.06103
20	S98T003399_S	4.4341	.07476	4.8753	2.8372
21	S98T003399_X	.10131	.07428	.67210	.08320
22	S98T003399_AX	198.13	.36144	187.93	199.05
23	CCV	5.0331	.00748	4.7955	4.9066
24	CCB	.00128	.00015	-.00035	.00102
25	S98T003389_L	.01339	-.00591	.03888	.00373
26	S98T003389	.02188	.00024	.02408	.01117
27	S98T003389_D	.02442	.00176	.02336	.01793
28	S98T003389_S	4.8504	.01044	4.8078	4.7685
29	S98T003389_X	.14182	.02254	.01091	.11677
30	S98T003389_AX	492.07	.74933	472.21	491.06
31	ICSA	.00252	.00752	-.00442	-.00053
32	ICSAB	.46786	.00769	.94459	-.00122
33	CCV	5.0311	.00766	4.8220	4.8939
34	CCB	.00109	.00000	.00058	.00109

Handwritten: 12-31-98

Handwritten: 12-31-98

LABCORE Data Entry Template for Worklist# 27676

Analyst: MX Scto Instrument: ICP01 h 12-30-98Book# 77B488Method: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysis

Worklist Comment: ICP BY-105 (DIRECT)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	SERDIL	S98T003384	0 D	@ICP-D01	LIQUID	
7	SAMPLE	S98T003384	0 D	@ICP-D01	LIQUID	98000437 BY-105 (2)
Analytes Requested: AG-D-01, AL-D-01, AS-D-01, B-D-01, BA-D-01, BE-D-01, BI-D-01, CA-D-01, CD-D-01, CE-D-01, CO-D-01, CR-D-01, CU-D-01, FE-D-01, K-D-01, LA-D-01, LI-D-01, MG-D-01, MN-D-01, MO-D-01, NA-D-01, ND-D-01, NI-D-01, P-D-01, PE-D-01, S-D-01, SE-D-01, SE-D-01, SI-D-01, SM-D-01, SR-D-01, TI-D-01, TL-D-01, U-D-01, V-D-01, ZN-D-01, ZR-D-01						
8	DUP	S98T003384	0 D	@ICP-D01	LIQUID	
9	SPK (1 ppm)	S98T003384	0 D	@ICP-D01	LIQUID	
10	SPK (10 ppm)	S98T003384	0 D	@ICP-QC	QC	
11	ICSAB		@ICP-QC	QC		
12	CCV		@ICP-QC	QC		
13	CCB		@ICP-QC	QC		

Data Entry Comments:

Uploaded & Validated by:
Paul M. Pung
1/15/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

12/28/98 13:12
A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27676

S Type	Sample#	R A	Test	Matrix	Group#	Project

Final page for worklist # 27676

Analyst Signature	Date
JK [Signature]	01-15-99
5987003384-L	025-15-28, 3F 3005
5987003384	025-15 601
5987003384-D	025-15 601
5987003384-G	025-15 601
5987003384-X	025-15-19 6010
5987003384-QK	025-15-19 6010

Analyst Signature	Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1977

Analysis Report

Summary

Fri 01-15-99 01:05:25 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990115B	ICP2	01/15/99	12:05	DKS	Q	CONC
2	ICB	990115B	ICP2	01/15/99	12:11	DKS	Q	CONC
3	LLS	990115B	ICP2	01/15/99	12:14	DKS	Q	CONC
4	ICSA	990115B	ICP2	01/15/99	12:16	DKS	Q	CONC
5	ICSAB	990115B	ICP2	01/15/99	12:19	DKS	Q	CONC
6	S98T003384_L	990115B	ICP2	01/15/99	12:23	DKS	S	CONC
7	S98T003384	990115B	ICP2	01/15/99	12:26	DKS	S	CONC
8	S98T003384_D	990115B	ICP2	01/15/99	12:29	DKS	S	CONC
9	S98T003384_S	990115B	ICP2	01/15/99	12:33	DKS	S	CONC
10	S98T003384_X	990115B	ICP2	01/15/99	12:38	DKS	S	CONC
11	S98T003384_AX	990115B	ICP2	01/15/99	12:41	DKS	S	CONC
12	ICSA	990115B	ICP2	01/15/99	12:51	DKS	Q	CONC
13	ICSAB	990115B	ICP2	01/15/99	12:54	DKS	Q	CONC
14	CCV	990115B	ICP2	01/15/99	12:58	DKS	Q	CONC
15	CCB	990115B	ICP2	01/15/99	13:02	DKS	Q	CONC

JX Ladd

BY-105

01-15-99

5987003384

Worklist # 27676

S98T003384

$$\text{Spk: } C_r = \frac{\left(\frac{2117.2}{601}\right) - \left(\frac{1516.1}{601}\right)}{1} \times 100 = 100.0\%$$

Post Spk:

$$Al = \frac{\left(\frac{96320}{6010}\right) - \left(\frac{39164}{6010}\right)}{10} \times 100 = 95.1\%$$

$$Na = \frac{\left(\frac{276530}{6010}\right) - \left(\frac{224290}{6010}\right)}{10} \times 100 = 86.9\%$$

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1978 TO 1981.

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9164	4.8838	4.9849	5.3050	4.8570	5.0754
2	ICB	.00000	.00255	.00703	.00773	-.00007	-.00013
3	LLS	.01524	.10267	.20787	.09708	.09937	.01048
4	ICSA	.00129	242.59	-.02598	-.00392	.00023	.00009
5	ICSAB	.94053	242.01	-.00075	-.00631	.45724	.47229
6	S98T003384_L	15.554	39098.	-16.459	34.841	-.29791	.40152
7	S98T003384	15.206	38282.	-26.791	38.303	-.00043	.48514
8	S98T003384_D	15.435	39106.	-29.840	39.178	-.00435	.40318
9	S98T003384_S	566.23	38595.	591.85	653.87	548.51	592.89
10	S98T003384_X	25.964	39164.	-8.2611	34.847	-.38097	.54119
11	S98T003384_AX	58613.	96320.	60185.	58967.	59281.	59257.
12	ICSA	.00195	236.29	-.00645	-.00319	.00025	.00019
13	ICSAB	.93866	238.74	-.00119	-.00494	.44998	.46742
14	CCV	4.9160	4.7755	4.9502	5.1949	4.7349	4.9960
15	CCB	.00058	.00328	.00026	.00097	-.00001	.00004

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0300	5.0502	5.0445	5.0354	4.9813	4.9455
2	ICB	.00203	-.02471	-.00030	.00194	-.00436	-.00075
3	LLS	.22508	.19886	.01028	.20686	.03832	.02020
4	ICSA	.02311	252.46	.00455	.00489	-.00189	-.00240
5	ICSAB	-.01818	253.13	.93387	.00296	.45202	.46013
6	S98T003384_L	15.741	-27.328	.13982	-5.8133	-7.5470	1567.2
7	S98T003384	10.276	20.573	1.1049	-1.9429	3.1012	1516.1
8	S98T003384_D	.46823	11.906	.82930	1.9335	.39535	1540.3
9	S98T003384_S	601.21	615.48	597.13	578.32	588.27	2117.2
10	S98T003384_X	45.407	-88.880	.52395	-29.170	-10.306	1565.0
11	S98T003384_AX	60324.	60679.	59615.	59694.	59320.	61455.
12	ICSA	.01734	256.71	.00478	.00907	-.00321	-.00171
13	ICSAB	-.01170	256.07	.93435	-.00137	.45374	.46331
14	CCV	4.9774	5.1404	5.0496	4.8986	4.9767	4.9601
15	CCB	.04013	-.01936	.00062	.00492	-.00289	-.00033

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.8157	.00642	4.7805	4.8621	5.0400	5.0412
2	ICB	.00040	-.00019	-.00199	.02025	-.00012	-.00010
3	LLS	.01804	.00015	.09614	Q.72323	.10168	.02132
4	ICSA	.00361	.00136	92.133	.28655	-.00398	.00204
5	ICSAB	.46551	.00238	91.917	.09498	-.00399	.97372
6	S98T003384_L	-2.1904	-.21634	23.906	4818.9	-2.5992	.30538
7	S98T003384	-.63247	.01827	26.824	4514.2	-.63678	.61114
8	S98T003384_D	-1.3148	-.22186	27.416	4495.0	-.29873	.36682
9	S98T003384_S	551.31	1.3007	583.76	5004.5	584.65	559.10
10	S98T003384_X	-1.0398	-.38619	12.126	5389.3	-2.4498	3.6653
11	S98T003384_AX	55562.	80.437	57213.	60742.	59339.	58448.
12	ICSA	.00394	-.01227	91.421	.29832	-.00363	.00346
13	ICSAB	.45435	-.00655	91.659	.20000	-.00449	.94555
14	CCV	4.6665	.00779	4.7367	4.7481	4.9076	4.8227
15	CCB	.00000	.00073	-.00241	.11272	.00034	.00081

Analysis Report

Averages

Fri 01-15-99 01:05:25 PM

page 3

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1896	4.8922	5.2439	5.0188	4.9275	4.9174
2	ICB	-.00104	-.00035	-.00155	-.01027	.00084	.00332
3	LLS	.21037	.02014	.10010	.20031	.20339	.03894
4	ICSA	247.95	-.00514	-.01753	191.39	.00296	-.00893
5	ICSAB	246.58	.43901	-.01583	191.77	-.00132	.91917
6	S98T003384_L	-7.3956	-.74408	31.649	224990.	1.4989	2.1773
7	S98T003384	-5.3796	-.10779	33.724	214170.	1.5036	-2.3598
8	S98T003384_D	-2.2462	-.18459	34.143	219800.	-2.0647	-1.6976
9	S98T003384_S	590.93	565.40	657.22	210620.	570.57	579.16
10	S98T003384_X	-37.625	-.94686	18.689	224290.	2.1502	-5.1134
11	S98T003384_AX	61522.	59406.	59541.	276530.	57861.	59536.
12	ICSA	241.95	-.00434	-.01702	182.13	.00039	-.00312
13	ICSAB	243.64	.43967	-.01567	186.60	.00072	.91258
14	CCV	5.0580	4.8758	5.2377	4.7656	4.7704	4.9336
15	CCB	.00103	.00005	-.00240	-.01524	.00103	.00145

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.2754	4.9622	5.0598	5.3049	4.8566	5.4074
2	ICB	.01325	-.00356	-.00422	.00000	-.00744	-.00580
3	LLS	.40125	.21709	.18523	.09832	.20468	Q.10926
4	ICSA	.03014	-.00041	-.03541	.00103	.03777	-.00005
5	ICSAB	.02865	.97847	-.04965	-.00706	.01330	-.00409
6	S98T003384_L	398.56	69.231	913.76	-4.5683	56.912	287.21
7	S98T003384	406.07	69.238	921.37	-3.4692	34.264	275.48
8	S98T003384_D	408.15	62.105	935.40	5.2907	46.315	282.20
9	S98T003384_S	1033.2	669.79	1526.1	627.22	663.97	942.32
10	S98T003384_X	498.53	68.696	842.33	8.8195	137.60	331.83
11	S98T003384_AX	60021.	60455.	59559.	59960.	58659.	62042.
12	ICSA	.04567	.01331	-.03863	.00079	.05304	.01669
13	ICSAB	.04019	.97642	-.05226	.01057	.03283	.00846
14	CCV	5.1887	4.9938	5.0442	5.2543	4.7915	5.3811
15	CCB	.02564	-.00785	-.00387	.00005	.02719	.00197

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8979	4.9231	.21433	5.3070	4.7525	9.8811
2	ICB	-.00204	-.00012	-.00138	.00000	-.01552	-.00700
3	LLS	.20347	.02021	.00458	.02047	.38405	.48420
4	ICSA	-.02149	.00181	-.00408	.00117	.03622	-.15774
5	ICSAB	-.01908	.00183	-.00662	.00094	.04479	-.16005
6	S98T003384_L	4.4835	.51224	-16.951	-.05960	44.560	3.6178
7	S98T003384	-.86439	.74107	-1.5368	.54373	1.8571	.53696
8	S98T003384_D	-2.9566	.71582	-1.2679	-.28360	4.9045	-2.1505
9	S98T003384_S	568.28	569.43	22.214	618.40	557.07	1141.8
10	S98T003384_X	16.995	.51180	-31.637	-1.4899	-73.908	33.135
11	S98T003384_AX	57467.	58991.	2567.3	59164.	58493.	112020.
12	ICSA	-.00738	.00150	.00170	.00191	-.00101	-.16577
13	ICSAB	-.01376	.00165	-.00177	.00143	.02354	-.17150
14	CCV	4.7482	4.8026	.21080	5.2416	4.6971	9.5533
15	CCB	.01049	.00004	-.00064	-.00046	.01188	.03090

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9975	.00786	4.7651	5.0771
2	ICB	-.00052	.00008	-.00068	.00011
3	LLS	.09967	.00042	.01881	.01898
4	ICSA	.00003	.00743	-.00467	-.00325
5	ICSAB	.46035	.00743	.92313	-.00317
6	S98T003384_L	-2.1843	-.00153	-2.4367	1.2306
7	S98T003384_	-.26482	-.04731	-.74064	.27974
8	S98T003384_D	-.91340	-.14034	-.49920	-.11173
9	S98T003384_S	586.31	1.1157	576.16	596.85
10	S98T003384_X	4.3469	1.3985	-5.2162	6.0989
11	S98T003384_AX	58899.	96.478	56412.	58699.
12	ICSA	.00386	.00821	-.00381	-.00149
13	ICSAB	.46178	.00790	.92929	-.00229
14	CCV	4.9591	.00878	4.8130	4.9948
15	CCB	.00212	.00077	-.00079	.00159

01-15-99

LABCORE Data Entry Template for Worklist# 27696

Analyst: JK Sato Instrument: ICP01 1 12-30-98 Book# 77848BMethod: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysisWorklist Comment: ICP BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTD1JA		@ICP-A01	SOLID		
7	PREPBLKTJA		@ICP-A01	SOLID		
8	SERDIL	S98T002928 0 A	@ICP-A01	SOLID		
9	SAMPLE	S98T002928 0 A	@ICP-A01	SOLID	98000490 BY-105 (2)	
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01, MG-A-01, MN-A-01, MO-A-01, NA-A-01, ND-A-01, NI-A-01, P-A-01, PB-A-01, S-A-01, SB-A-01, SE-A-01, SI-A-01, SM-A-01, SR-A-01, TI-A-01, TL-A-01, U-A-01, V-A-01, ZN-A-01, ZR-A-01						
10	DUP	S98T002928 0 A	@ICP-A01	SOLID		
11	SPK-PREDIG	S98T002928 0 A	@ICP-A01	SOLID		
12	CCV	<u>Sample 1 (10 ppn) 12-30-98</u>	@ICP-QC	QC		
13	CCB		@ICP-QC	QC		
14	SAMPLE	S98T002929 0 A	@ICP-A01	SOLID	98000490 BY-105 (2)	
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01						

Data Entry Comments:

uploaded 1-6-99
John W. Swell

Validated by:

Paul M. Yang
01/06/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

12/29/98 08:24

A-0004-1

HNF-1663 REV. 0

Page: 2

LABCORE Data Entry Template for Worklist# 27696

S Type	Sample#	R A	Test	Matrix	Group#	Project			
			MG-A-01	MN-A-01	MO-A-01	NA-A-01	ND-A-01	NI-A-01	
			P-A-01	PB-A-01	S-A-01	SB-A-01	SE-A-01	SI-A-01	
			SM-A-01	SR-A-01	TI-A-01	TL-A-01	U-A-01	V-A-01	
			ZN-A-01	ZR-A-01					
15 DUP	S98T002929	0 A	@ICP-A01	SOLID					
16 SERDIL	S98T002930	0 A	@ICP-A01	SOLID					
17 SAMPLE	S98T002930	0 A	@ICP-A01	SOLID	98000490	BY-105 (2)			
	Analytes Requested:								
			AG-A-01	AL-A-01	AS-A-01	B-A-01	BA-A-01		
			BE-A-01	BI-A-01	CA-A-01	CD-A-01	CE-A-01	CO-A-01	
			CR-A-01	CU-A-01	FE-A-01	K-A-01	LA-A-01	LI-A-01	
			MG-A-01	MN-A-01	MO-A-01	NA-A-01	ND-A-01	NI-A-01	
			P-A-01	PB-A-01	S-A-01	SB-A-01	SE-A-01	SI-A-01	
			SM-A-01	SR-A-01	TI-A-01	TL-A-01	U-A-01	V-A-01	
			ZN-A-01	ZR-A-01					
18 DUP	S98T002930	0 A	@ICP-A01	SOLID					
19 SPK-PREDIG	S98T002930	0 A	@ICP-A01	SOLID					
20 ICSA	S98T002930	0 A	@ICP-QC	QC					
21 ICSAB	S98T002930	0 A	@ICP-QC	QC					
22 CCV	S98T002930	0 A	@ICP-QC	QC					
23 CCB	S98T002930	0 A	@ICP-QC	QC					

Final page for worklist # 27696

Analyst Signature Date
 Prep 1/1/99 Direct 1/1/99
 Prep 1/1/99 Direct 1/1/99
 S98T002929-L 1-8-98 1/1/99
 S98T002929 1-8-98 5
 S98T002929-D 1-8-98 5
 S98T002929-S 1-8-98 5
 S98T002929-X 1-8-98 50
 S98T002929-QX 1-8-98 50
 S98T002929 1-8-98 5
 S98T002929-D 1-8-98 5
 Data Entry Comments:

Analyst Signature Date
 S98T002930-L 1-8-98 1/1/99
 S98T002930 1-8-98 5
 S98T002930-D 1-8-98 5
 S98T002930-S 1-8-98 5
 S98T002930-X 1-8-98 50
 S98T002930-QX 1-8-98 50

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1983

S98 T00 2928

$$Al = 9.0347 \frac{\mu g}{ml} \times \frac{1}{4.904 \frac{g}{L} \times \frac{1L}{100ml}} = 1.842 \pm 3 \frac{\mu g}{g}$$

$$\text{Pre-dig Spk: } Al = \frac{15.064 - 9.0347}{5} \times 100 = 121.0\%$$

Post Spk:

$$Al = \frac{\left(\frac{492.99}{50}\right) - \left(\frac{9.4263}{50}\right)}{10} \times 100 = 96.7\%$$

$$Na = \frac{\left(\frac{1774.3}{50}\right) - \left(\frac{1315.5}{50}\right)}{10} \times 100 = 91.8\%$$

S98 T00 2930

$$Al = 76.921 \frac{\mu g}{ml} \times \frac{1}{5.144 \frac{g}{L} \times \frac{1L}{100ml}} = 1.502 \pm 4 \frac{\mu g}{g}$$

$$\text{Pre-dig Spk: } Al = \frac{81.135 - 76.921}{5} \times 100 = 84.3\%$$

Post Spk:

$$Fe = \frac{\left(\frac{532.91}{50}\right) - \left(\frac{55.088}{50}\right)}{10} \times 100 = 95.6\%$$

$$Na = \frac{\left(\frac{1696.0}{50}\right) - \left(\frac{1178.6}{50}\right)}{10} \times 100 = 103.5\%$$

$$U = \frac{\left(\frac{1032.6}{50}\right) - \left(\frac{84.539}{50}\right)}{20} \times 100 = 94.8\%$$

Analysis Report

Summary

Wed 01-06-99 12:36:39 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990106A	ICP2	01/06/99	08:35	DKS	Q	CONC
2	ICB	990106A	ICP2	01/06/99	08:38	DKS	Q	CONC
3	LLS	990106A	ICP2	01/06/99	08:41	DKS	Q	CONC
4	ICSA	990106A	ICP2	01/06/99	08:44	DKS	Q	CONC
5	ICSAB	990106A	ICP2	01/06/99	08:47	DKS	Q	CONC
6	PREPSTDTJA	990106A	ICP2	01/06/99	08:50	DKS	Q	CONC
7	PREPBLKTJA	990106A	ICP2	01/06/99	09:13	DKS	Q	CONC
8	S98T002928_L	990106A	ICP2	01/06/99	09:17	DKS	S	CONC
9	S98T002928	990106A	ICP2	01/06/99	09:20	DKS	S	CONC
10	S98T002928_D	990106A	ICP2	01/06/99	09:23	DKS	S	CONC
11	S98T002928_S	990106A	ICP2	01/06/99	09:26	DKS	S	CONC
12	S98T002928_X	990106A	ICP2	01/06/99	09:29	DKS	S	CONC
13	S98T002928_AX	990106A	ICP2	01/06/99	09:33	DKS	S	CONC
14	CCV	990106A	ICP2	01/06/99	10:09	DKS	Q	CONC
15	CCB	990106A	ICP2	01/06/99	10:12	DKS	Q	CONC
16	S98T002929	990106A	ICP2	01/06/99	10:16	DKS	S	CONC
17	S98T002929_D	990106A	ICP2	01/06/99	10:20	DKS	S	CONC
18	S98T002930_L	990106A	ICP2	01/06/99	10:24	DKS	S	CONC
19	S98T002930	990106A	ICP2	01/06/99	10:27	DKS	S	CONC
20	S98T002930_D	990106A	ICP2	01/06/99	10:30	DKS	S	CONC
21	S98T002930_S	990106A	ICP2	01/06/99	10:33	DKS	S	CONC
22	S98T002930_X	990106A	ICP2	01/06/99	10:37	DKS	S	CONC
23	S98T002930_AX	990106A	ICP2	01/06/99	10:40	DKS	S	CONC
24	ICSA	990106A	ICP2	01/06/99	12:18	DKS	Q	CONC
25	ICSAB	990106A	ICP2	01/06/99	12:21	DKS	Q	CONC
26	CCV	990106A	ICP2	01/06/99	12:24	DKS	Q	CONC
27	CCB	990106A	ICP2	01/06/99	12:27	DKS	Q	CONC

JK Ltd

01-06-99

Worklist # 27686

BY-105

5987002928

5987002929

5987002930

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/REVIEWED THE CALIBRATION/ANALYSIS ON PAGES 1985 TO 1989.

Analysis Report

Averages

Wed 01-06-99 12:36:39 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9123	4.8243	4.9825	5.3190	4.7905	5.0403
2	ICB	.00234	-.00063	-.01165	.01514	-.00013	.00004
3	LLS	.02217	.09529	.19254	.10942	.09780	.01043
4	ICSA	.00268	238.39	.03769	-.00146	.00001	.00013
5	ICSAB	.94748	239.68	.03248	.00613	.45857	.47702
6	PREPSTD TJ A	.84961	4.7024	4.5301	5.3820	4.5918	5.0252
7	PREPBLKTJ A	.00002	Q.12319	-.00624	Q.60842	.00039	-.00018
8	S98T002928_L	.07956	9.4449	-.24468	.84266	-.00320	.00225
9	S98T002928	.08907	9.0347	-.06396	.69843	.00218	.00000
10	S98T002928_D	.09777	9.7378	-.05045	.79853	.00212	.00000
11	S98T002928_S	1.0315	15.064	4.9738	5.5901	4.6605	5.0067
12	S98T002928_X	.13901	9.4263	-.74537	.92793	-.00528	.00461
13	S98T002928_AX	129.59	492.99	503.55	493.30	497.16	497.35
14	CCV	4.9796	4.8903	5.0545	5.3907	4.8944	5.1212
15	CCB	.00887	.00023	-.01490	.00342	-.00016	.00000
16	S98T002929	.09933	42.047	-.06975	.63740	.00177	-.00023
17	S98T002929_D	.08582	41.272	-.10449	.76675	.00086	.00047
18	S98T002930_L	.12965	79.020	-.11426	.74363	.61818	.00129
19	S98T002930	.10254	76.921	-.13724	.73011	.60732	.00188
20	S98T002930_D	.11155	71.202	-.09677	.58412	.49651	.00693
21	S98T002930_S	1.0434	81.135	5.0591	5.5677	5.2346	5.1108
22	S98T002930_X	.16967	79.357	-.44003	.84151	.61828	-.00218
23	S98T002930_AX	118.74	567.34	506.88	496.19	503.68	501.93
24	ICSA	.00635	236.15	.04796	-.00526	.00001	.00050
25	ICSAB	.91871	238.73	.02755	-.00247	.46953	.48816
26	CCV	4.7404	4.7918	4.9969	5.3124	4.9088	5.1722
27	CCB	.00243	.01832	-.00447	.00779	-.00032	.00027

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9308	5.1148	5.0556	4.9560	5.0094	4.9468
2	ICB	-.00412	.00232	.00059	.00104	-.00088	.00082
3	LLS	.17357	.22352	.01052	.20217	.03905	.02072
4	ICSA	-.04026	251.23	.00635	.00374	.00024	-.00138
5	ICSAB	.00744	253.44	.94721	.00828	.46038	.46811
6	PREPSTD TJ A	4.4356	4.8732	4.5939	4.7583	4.6204	4.6360
7	PREPBLKTJ A	-.03785	Q.25862	-.00046	-.00144	.00136	.00257
8	S98T002928_L	-.28967	.59151	.01241	.06960	.00360	.90595
9	S98T002928	-.10716	.58380	.00762	.00217	.00322	.87639
10	S98T002928_D	.01719	.54789	.00513	-.00327	-.01212	.90913
11	S98T002928_S	4.9345	5.5033	5.0666	4.9473	4.9909	5.8673
12	S98T002928_X	-.68692	.59468	.01468	.08514	.00634	.89797
13	S98T002928_AX	502.97	507.78	498.85	499.85	497.52	501.12
14	CCV	5.0443	5.1965	5.1198	5.0475	5.0673	5.0237
15	CCB	-.02014	.00463	.00072	.00038	-.00011	.00039
16	S98T002929	-.04213	1.0733	.00305	-.01002	.00186	1.8736
17	S98T002929_D	-.15628	.52399	.00192	-.06805	.00121	1.8816
18	S98T002930_L	4.1251	140.85	-.00516	.09959	.02218	3.2867
19	S98T002930	4.8473	136.44	.00914	.01536	.02531	3.1646
20	S98T002930_D	3.2552	116.67	.01849	.05938	.03587	3.1583
21	S98T002930_S	8.1956	132.64	5.1453	5.0206	5.0246	8.3789
22	S98T002930_X	5.8279	142.55	.04732	.59603	.06738	3.3427

Analysis Report

Averages

Wed 01-06-99 12:36:39 PM

page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
23	S98T002930_AX	512.34	642.91	499.42	503.00	498.82	503.81
24	ICSA	.00596	242.51	.00534	.00850	.00059	-.00159
25	ICSAB	-.02946	244.35	.93451	-.00091	.45565	.45949
26	CCV	4.9520	4.9369	5.0039	5.0691	4.9166	4.8532
27	CCB	-.00455	-.00011	.00100	.00324	-.00080	-.00056

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7610	.00628	4.7720	4.9103	4.9695	4.9418
2	ICB	.00037	.00034	.00117	.13427	.00071	.00032
3	LLS	.01696	.00045	.09729	.51377	.10113	.02079
4	ICSA	.00288	.00846	91.921	.08749	-.00296	.00261
5	ICSAB	.46121	.00929	92.372	.02147	-.00348	.97265
6	PREPSTDITJA	4.3486	.00547	4.4502	4.4664	4.7408	4.7227
7	PREPBLKTJA	Q.01037	-.00158	.02845	.01642	-.00067	-.00119
8	S98T002928_L	-.00233	.03112	.23771	3.8598	.02533	.02990
9	S98T002928	.00790	.00507	.25750	1.9191	.00249	.00434
10	S98T002928_D	.00630	.00110	.26872	1.9510	.00604	.00162
11	S98T002928_S	4.5846	.00799	4.9661	7.0872	4.8682	4.7701
12	S98T002928_X	.02678	.00653	.29305	6.4243	.01039	.02178
13	S98T002928_AX	465.69	.59284	478.66	492.16	496.19	487.67
14	CCV	4.8263	.00614	4.8373	4.9152	5.0628	5.0306
15	CCB	.00048	.00094	.00083	.14278	.00084	.00054
16	S98T002929	.00677	.00256	.44083	5.2848	-.00073	.00108
17	S98T002929_D	-.00013	-.00012	.28830	3.7707	-.01081	.00000
18	S98T002930_L	-.23213	.17036	54.976	6.8314	.03244	.02821
19	S98T002930	-.23115	.12983	52.634	4.9329	.02987	.01090
20	S98T002930_D	-.12753	.11845	35.690	5.9818	.02976	.01739
21	S98T002930_S	4.2466	.13176	38.577	9.8313	4.9593	4.8579
22	S98T002930_X	-.20546	.22543	55.088	9.2243	.14275	.06629
23	S98T002930_AX	470.62	.73699	532.91	489.21	501.25	494.72
24	ICSA	.00397	.02179	90.863	.10200	-.00303	.00087
25	ICSAB	.47514	.02378	91.524	-.02367	-.00424	.99173
26	CCV	4.8666	.00391	4.7163	4.8605	5.1121	4.9984
27	CCB	.00120	-.00226	.00194	-.01389	.00122	-.00141

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1393	4.8703	5.2522	4.9153	4.8452	4.8961
2	ICB	-.00186	.00011	.00068	-.00297	.00155	-.00094
3	LLS	.20564	.02040	.09922	.20563	.19745	.03997
4	ICSA	246.26	-.00473	-.01791	190.15	.00337	-.01629
5	ICSAB	246.78	.44363	-.01862	189.12	.00098	.91050
6	PREPSTDITJA	4.7036	4.5312	4.6343	5.8292	4.6748	4.5212
7	PREPBLKTJA	.07272	.00042	-.00223	Q.98673	.00117	.00582
8	S98T002928_L	.14306	.02157	-.03295	1301.2	-.03690	.11815
9	S98T002928	.09823	.01367	-.00827	1236.0	.01176	.15564
10	S98T002928_D	.07970	.01218	-.01088	1341.6	-.00418	.15096
11	S98T002928_S	5.1363	4.7754	5.0103	1314.8	4.7675	5.0345
12	S98T002928_X	.24438	.01349	-.14357	1315.5	.01485	.17185
13	S98T002928_AX	510.79	497.94	499.04	1774.3	482.10	496.33
14	CCV	5.1687	4.9421	5.3224	4.9836	4.9337	4.9521

Analysis Report

Averages

Wed 01-06-99 12:36:39 PM

page 4

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
15	CCB	-.00053	.00026	.00035	-.00074	.00082	.00000
16	S98T002929	.12362	.01701	.02283	1234.2	.01026	.10308
17	S98T002929_D	.00377	.01427	.01188	1191.8	-.01006	.12909
18	S98T002930_L	3.8158	.78069	-.01496	1194.3	.22067	27.992
19	S98T002930	3.6250	.74840	.00564	1154.2	.20591	26.680
20	S98T002930_D	2.6465	.50673	.02053	1121.7	.16707	16.205
21	S98T002930_S	7.7190	5.3118	5.0452	1056.2	4.9963	19.806
22	S98T002930_X	4.1628	.80241	-.02707	1178.6	.30273	28.533
23	S98T002930_AX	519.03	499.24	500.69	1696.0	487.80	525.03
24	ICSA	242.52	-.00552	-.01456	189.39	.00361	-.00426
25	ICSAB	244.57	.43444	-.01680	190.62	.00435	.91616
26	CCV	5.0627	4.7919	5.1992	4.9363	4.9691	4.8184
27	CCB	.00167	-.00055	.00045	.03907	-.00088	-.00127

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.2447	4.9855	5.1061	5.3064	4.7961	5.4141
2	ICB	.00104	.00867	-.00012	.00493	.00915	.02582
3	LLS	.40105	.20679	.20032	.10163	.21436	.12487
4	ICSA	.04587	.04560	-.04335	.00879	.06161	.00523
5	ICSAB	.03652	1.0030	-.03971	.00382	.06476	.00096
6	PREPBLDTJA	4.4735	4.4345	4.2987	4.4080	4.5118	9.1219
7	PREPBLKTJA	.03167	.00567	Q.09546	.00323	.02721	Q2.1180
8	S98T002928_L	1.5740	.15821	5.3000	-.02590	.70121	2.0704
9	S98T002928	1.4615	.06956	5.1733	.03377	.17149	1.6177
10	S98T002928_D	1.9314	.03255	5.6458	-.00240	.07041	1.1278
11	S98T002928_S	6.9593	5.0332	11.853	5.0349	4.9378	8.9760
12	S98T002928_X	2.0140	.17892	5.3537	.25493	.85097	2.1738
13	S98T002928_AX	500.74	502.61	494.42	502.67	487.19	493.56
14	CCV	5.3087	5.0461	5.1150	5.3632	4.8893	5.4192
15	CCB	.01607	.00218	-.00361	.00386	.01684	.00276
16	S98T002929	2.0429	.12078	5.2389	.00033	.18226	1.1920
17	S98T002929_D	2.2234	.07818	4.6359	.00992	.13023	1.1122
18	S98T002930_L	52.683	1.6691	43.947	.10901	.48731	17.417
19	S98T002930	50.920	1.4805	42.990	.01290	.13590	16.839
20	S98T002930_D	36.815	1.1224	13.694	.02642	.15615	21.520
21	S98T002930_S	45.740	6.1315	16.846	5.0173	5.0145	27.486
22	S98T002930_X	52.703	1.8862	44.172	.44960	1.3802	18.290
23	S98T002930_AX	557.87	506.06	535.70	503.84	491.35	510.35
24	ICSA	.04463	.03417	-.03556	-.00438	.09284	-.01430
25	ICSAB	.06477	.98034	-.03962	.00381	.08632	-.01627
26	CCV	5.1508	4.8440	4.8779	5.2608	4.8457	5.1881
27	CCB	.03671	.00361	-.00040	.00484	.05579	-.01053

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8143	4.8538	.17549	5.2863	4.7370	9.7167
2	ICB	.00366	.00000	.00179	.00000	-.00764	.01286
3	LLS	.19708	.01995	.00451	.02012	.40150	.48456
4	ICSA	-.01262	.00181	-.00157	.00206	-.00628	.02063
5	ICSAB	-.01499	.00177	.00128	.00160	-.01807	.02233
6	PREPSTDTJA	4.6422	4.7000	.16204	4.6905	4.3859	9.0632

HNF-1663 REV. 0

Analysis Report

Averages

Wed 01-06-99 12:36:39 PM

page 5

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
7	PREPBLKTJA	-.02036	.00026	.00278	.00094	-.00179	-.07027
8	S98T002928_L	.22702	.03884	-.09274	-.00074	-.08416	1.4092
9	S98T002928	.04609	.03448	.00673	.00457	.04454	.36189
10	S98T002928_D	.00018	.03450	.00307	.00108	.02827	.17339
11	S98T002928_S	4.7474	4.8511	.19103	4.9367	4.6841	9.4246
12	S98T002928_X	.01558	.03560	.20111	.01229	-.23746	.30384
13	S98T002928_AX	480.93	490.74	17.548	494.67	486.50	939.09
14	CCV	4.9076	4.9391	.18247	5.3655	4.8191	9.8879
15	CCB	.00661	.00004	-.00083	.00092	-.00693	.02770
16	S98T002929	.00301	.02313	-.00241	.01046	-.06135	.05711
17	S98T002929_D	-.01102	.02515	-.01788	-.00010	.01040	.04385
18	S98T002930_L	.20400	20.659	1.1078	.30021	-.38961	83.102
19	S98T002930	.03492	20.213	1.1728	.29445	-.00922	80.681
20	S98T002930_D	.06163	17.010	.70837	.27628	-.06716	51.283
21	S98T002930_S	4.8396	22.510	.85381	5.2509	4.6809	59.475
22	S98T002930_X	.87423	20.527	1.3776	.32480	.12417	84.539
23	S98T002930_AX	486.88	516.03	19.556	498.35	495.07	1032.6
24	ICSA	-.03812	.00194	.00526	.00114	-.00626	-.03376
25	ICSAB	-.04172	.00196	.00648	.00161	.02076	-.03957
26	CCV	4.9936	4.9701	.20087	5.3346	4.8106	10.023
27	CCB	-.02620	-.00043	.01203	-.00017	-.01265	-.09338

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9869	.00762	4.8022	5.0421
2	ICB	-.00020	.00030	-.00001	.00017
3	LLS	.10036	.00049	.01961	.01926
4	ICSA	.00123	.00743	-.00381	-.00278
5	ICSAB	.46788	.00751	.93483	-.00316
6	PREPSTDTJA	4.6071	.00617	4.1648	4.4806
7	PREPBLKTJA	-.00404	-.00114	Q.01417	-.00045
8	S98T002928_L	.05634	.01703	.02865	.05423
9	S98T002928	.01461	.00337	.02955	.01211
10	S98T002928_D	.00232	.00071	.02393	.00611
11	S98T002928_S	4.8784	.00970	4.8075	4.7088
12	S98T002928_X	.00807	.01554	-.01489	.01735
13	S98T002928_AX	493.23	.76865	470.29	493.46
14	CCV	5.0639	.00778	4.8554	5.1319
15	CCB	.00139	.00053	-.00026	.00132
16	S98T002929	.00335	.00036	.02019	.00655
17	S98T002929_D	.00168	-.00041	.01691	.00616
18	S98T002930_L	.08168	.04109	.26284	.06188
19	S98T002930	.04039	.02684	.24939	.02053
20	S98T002930_D	.04228	.02444	.44717	.02528
21	S98T002930_S	4.9701	.03271	5.0337	4.7073
22	S98T002930_X	.20922	.07771	.26266	.11291
23	S98T002930_AX	495.29	.77594	469.31	496.26
24	ICSA	-.00367	.00597	-.00402	-.00695
25	ICSAB	.46017	.00589	.89205	-.00720
26	CCV	4.9713	.00608	4.5563	5.0947
27	CCB	-.00510	-.00144	.00008	-.00436

JK L&D
01-06-99

1989

LABCORE Data Entry Template for Worklist# 27738

Analyst: JA Sc70 Instrument: ICP011 BL-4-99Book# 77848BMethod: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysisWorklist Comment: ICP BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTDITJA		@ICP-A01	SOLID		
7	PREPBLKTJA		@ICP-A01	SOLID		
8	SAMPLE	S98T002607 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01, MG-A-01, MN-A-01, MO-A-01, NA-A-01, ND-A-01, NI-A-01, P-A-01, PB-A-01, S-A-01, SB-A-01, SE-A-01, SI-A-01, SM-A-01, SR-A-01, TI-A-01, TL-A-01, U-A-01, V-A-01, ZN-A-01, ZR-A-01						
9	DUP	S98T002607 0 A	@ICP-A01	SOLID		
10	CCV		@ICP-QC	QC		
11	CCB		@ICP-QC	QC		
12	SERDIL	S98T002608 0 A	@ICP-A01	SOLID		
13	SAMPLE	S98T002608 0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-A-01, AL-A-01, AS-A-01, B-A-01, BA-A-01, BE-A-01, BI-A-01, CA-A-01, CD-A-01, CE-A-01, CO-A-01, CR-A-01, CU-A-01, FE-A-01, K-A-01, LA-A-01, LI-A-01, MG-A-01, MN-A-01, MO-A-01, NA-A-01, ND-A-01, NI-A-01, P-A-01, PB-A-01, S-A-01, SB-A-01, SE-A-01, SI-A-01						

Data Entry Comments:

Uploaded & Validated by:

Sally M. Parong

1/19/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27738

S Type	Sample#	R A	Test	Matrix	Group#	Project
			SM-A-01 ,	SR-A-01 ,	TI-A-01 ,	TL-A-01 , U-A-01 , V-A-01 ,
			ZN-A-01 ,	ZR-A-01		
14 DUP	S98T002608	0 A	@ICP-A01	SOLID		
15 SPK-PREDIG	S98T002608	0 A	@ICP-A01	SOLID		
16 SAMPLE	S98T002609	0 A	@ICP-A01	SOLID	98000436	BY-105 (2)
	Analytes Requested:					
			AG-A-01 ,	AL-A-01 ,	AS-A-01 ,	B-A-01 , BA-A-01 ,
			BE-A-01 ,	BI-A-01 ,	CA-A-01 ,	CD-A-01 , CE-A-01 , CO-A-01 ,
			CR-A-01 ,	CU-A-01 ,	FE-A-01 ,	K-A-01 , LA-A-01 , LI-A-01 ,
			MG-A-01 ,	MN-A-01 ,	MO-A-01 ,	NA-A-01 , ND-A-01 , NI-A-01 ,
			P-A-01 ,	PB-A-01 ,	S-A-01 ,	SB-A-01 , SE-A-01 , SI-A-01 ,
			SM-A-01 ,	SR-A-01 ,	TI-A-01 ,	TL-A-01 , U-A-01 , V-A-01 ,
			ZN-A-01 ,	ZR-A-01		
17 DUP	S98T002609	0 A	@ICP-A01	SOLID		
18 ICESA			@ICP-QC	QC		
19 ICSAB			@ICP-QC	QC		
20 CCV			@ICP-QC	QC		
21 CCB			@ICP-QC	QC		

Final page for worklist # 27738

JK *[Signature]* 01-14-99

Analyst Signature	Date
Prep STD TJR	Direct, DF 1
Prep ILC TJR	Direct, 1
5987001607	2-6, DF 4
5987001607-D	2-6, 4
5987002608-L	4-4-2-8, DF 10
5987002608	4-4, 2
5987001608-D	4-4, 2
5987002608-S	4-4, 2
5987001608-X	4-4-1-9, 20
5987001608-4X	4-4-1-9, 20

Data Entry Comments:

Analyst Signature	Date
5987001609	4-4, DF 2
5987001609-D	4-4, 2

$$Cr = 3.6125 \frac{\mu g}{ml} \times \frac{1}{5.0560 \frac{g}{L} \times \frac{1L}{1000 ml}} = 7.14e+2 \frac{\mu g}{g}$$

Pre-dig Spk:

$$Cr = \frac{7.4790 - 3.6125}{5} \times 100 = 77.3\%$$

Post Spk:

$$Cr = \frac{\left(\frac{199.72}{20}\right) - \left(\frac{3.9933}{20}\right)}{10} \times 100 = 97.9\%$$

$$Fe = \frac{\left(\frac{335.28}{20}\right) - \left(\frac{149.02}{20}\right)}{10} \times 100 = 93.1\%$$

$$Na = \frac{\left(\frac{899.39}{20}\right) - \left(\frac{637.95}{20}\right)}{10} \times 100 = 100.7\%$$

$$Si = \frac{\left(\frac{215.90}{20}\right) - \left(\frac{7.0702}{20}\right)}{10} \times 100 = 104.4\%$$

$$U = \frac{\left(\frac{578.01}{20}\right) - \left(\frac{190.60}{20}\right)}{20} \times 100 = 96.9\%$$

$$Zn = \frac{\left(\frac{183.39}{20}\right) - \left(\frac{0.68641}{20}\right)}{10} \times 100 = 91.35\%$$

$$Zr = \frac{\left(\frac{199.50}{20}\right) - \left(\frac{0.01300}{20}\right)}{10} \times 100 = 99.7\%$$

$$Cu = \frac{\left(\frac{191.03}{20}\right) - 0}{10} \times 100 = 95.5\%$$

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990119B	ICP2	01/19/99	10:20	DKS	Q	CONC
2	ICB	990119B	ICP2	01/19/99	10:23	DKS	Q	CONC
3	LLS	990119B	ICP2	01/19/99	10:26	DKS	Q	CONC
4	ICSA	990119B	ICP2	01/19/99	10:28	DKS	Q	CONC
5	ICSAB	990119B	ICP2	01/19/99	10:32	DKS	Q	CONC
6	PREPSTDITJA	990119B	ICP2	01/19/99	10:35	DKS	Q	CONC
7	PREPBLKTJA	990119B	ICP2	01/19/99	11:57	DKS	Q	CONC
8	S98T002607	990119B	ICP2	01/19/99	12:01	DKS	S	CONC
9	S98T002607_D	990119B	ICP2	01/19/99	12:04	DKS	S	CONC
10	CCV	990119B	ICP2	01/19/99	12:10	DKS	Q	CONC
11	CCB	990119B	ICP2	01/19/99	12:13	DKS	Q	CONC
12	S98T002608_L	990119B	ICP2	01/19/99	12:16	DKS	S	CONC
13	S98T002608	990119B	ICP2	01/19/99	12:19	DKS	S	CONC
14	S98T002608_D	990119B	ICP2	01/19/99	12:23	DKS	S	CONC
15	S98T002608_S	990119B	ICP2	01/19/99	12:26	DKS	S	CONC
16	S98T002608_X	990119B	ICP2	01/19/99	12:33	DKS	S	CONC
17	S98T002608_AX	990119B	ICP2	01/19/99	12:36	DKS	S	CONC
18	S98T002609	990119B	ICP2	01/19/99	12:58	DKS	S	CONC
19	S98T002609_D	990119B	ICP2	01/19/99	13:02	DKS	S	CONC
20	ICSA	990119B	ICP2	01/19/99	13:06	DKS	Q	CONC
21	ICSAB	990119B	ICP2	01/19/99	13:09	DKS	Q	CONC
22	CCV	990119B	ICP2	01/19/99	13:13	DKS	Q	CONC
23	CCB	990119B	ICP2	01/19/99	13:16	DKS	Q	CONC

JK *[Signature]*

BY 105

01-19-99

S98T002607

Worklist # 27738

S98T002608

S98T002609

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 1993 TO 1997.

Analysis Report

Averages

Tue 01-19-99 01:19:19 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.8876	4.9590	5.0621	5.3701	4.9851	5.1751
2	ICB	.00121	.00480	-.00281	.00343	.00000	.00018
3	LLS	Q.01482	.11225	.20687	.10681	.10303	.01080
4	ICSA	.00179	248.97	-.04206	-.00430	.00028	.00013
5	ICSAB	.94814	247.54	-.02923	-.00510	.47025	.48564
6	PREPSTD TJA	.82025	4.6760	4.5304	5.2055	4.7037	5.0880
7	PREPBLKTJA	.00026	Q.11673	.00188	Q.54220	.00069	.00004
8	S98T002607	.09946	106.16	-.00224	.74800	.00317	.00073
9	S98T002607_D	.09568	89.157	-.01522	.38978	.00318	.00055
10	CCV	4.9119	4.9159	5.0396	5.3174	4.8833	5.1516
11	CCB	.00122	.00107	-.00019	.00589	.00011	.00023
12	S98T002608_L	.10448	131.05	-.10009	.31126	2.4417	.00528
13	S98T002608	.09830	125.08	-.11584	.30233	2.3132	.00191
14	S98T002608_D	.09273	119.67	-.05910	.18957	2.0568	.00183
15	S98T002608_S	.85580	130.31	4.1289	4.5661	6.2315	4.3123
16	S98T002608_X	.11491	131.75	-.15558	.29681	2.4443	.00623
17	S98T002608_AX	143.55	327.26	200.60	198.86	204.21	200.47
18	S98T002609	.12115	101.94	-.11207	.39333	2.4615	.01769
19	S98T002609_D	.10229	98.673	-.07341	.43921	2.3184	.00183
20	ICSA	.00235	240.19	.01462	-.00264	.00035	.00019
21	ICSAB	.93860	239.00	-.01323	-.01152	.44658	.47047
22	CCV	4.8851	4.7712	4.9197	5.1456	4.6851	4.9979
23	CCB	.00139	.00753	.00221	.00443	.00010	.00013

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0961	4.9948	5.0305	5.1589	4.9653	4.9118
2	ICB	-.00992	-.00610	.00050	.00289	-.00124	.00009
3	LLS	.24955	.21702	.01011	.21596	.04000	.02127
4	ICSA	.00120	249.78	.00542	-.00061	-.00018	-.00217
5	ICSAB	-.02901	252.40	.94246	-.00142	.45500	.46385
6	PREPSTD TJA	4.4090	4.5168	4.4176	4.8786	4.4635	4.4762
7	PREPBLKTJA	.02999	Q.14553	.00057	.00345	-.00101	.00207
8	S98T002607	.09837	.26096	.01306	.01575	.00363	.91866
9	S98T002607_D	-.02385	.20972	.01297	.05446	.00633	.80260
10	CCV	4.9899	5.0821	5.0671	5.0821	4.9959	4.9508
11	CCB	.03100	-.00660	-.00082	.00093	-.00123	-.00044
12	S98T002608_L	15.431	155.38	-.00096	-.09967	.06847	3.9713
13	S98T002608	14.574	143.12	.00942	-.04582	.08764	3.6125
14	S98T002608_D	12.998	137.97	.00561	-.04635	.07550	3.3422
15	S98T002608_S	18.178	142.70	3.9755	4.0592	3.5330	7.4790
16	S98T002608_X	15.352	156.68	.00608	-.02699	.05430	3.9933
17	S98T002608_AX	217.61	349.66	195.56	202.04	195.49	199.72
18	S98T002609	17.858	107.78	.02141	-.00875	.10890	1.9022
19	S98T002609_D	17.303	105.14	.00368	-.01922	.08502	1.8116
20	ICSA	-.00960	256.74	.00568	.00454	-.00071	-.00149
21	ICSAB	-.01485	255.21	.93344	.00444	.54575	.46026
22	CCV	4.9517	5.1434	5.0266	4.8680	4.9584	4.9399
23	CCB	-.01307	-.00400	.00056	.00568	-.00201	.00002

Analysis Report

Averages

Tue 01-19-99 01:19:19 PM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.9807	.00329	4.7933	5.0400	5.1884	5.2534
2	ICB	.00052	-.00203	-.00051	-.13874	.00006	-.00163
3	LLS	.01916	-.00187	.10003	.54121	.10723	.02103
4	ICSA	.00434	.00343	93.088	.01013	-.00376	.00021
5	ICSAB	.48219	-.00055	93.247	-.09710	-.00423	1.0095
6	PREPSTD TJJA	4.5106	.00203	4.3832	4.5867	4.8972	4.9914
7	PREPBLKTJA	.00352	-.00377	.02438	-.13012	.00003	-.00297
8	S98T002607	.02590	-.00782	.35628	1.2830	.00443	-.00492
9	S98T002607_D	.02211	-.00754	.30861	1.8031	.00628	-.00451
10	CCV	4.8829	.00425	4.8135	4.6997	5.1026	5.1070
11	CCB	.00101	-.00132	-.00058	.01647	-.00041	-.00112
12	S98T002608_L	-.55132	.13509	147.76	4.7658	.08278	.00711
13	S98T002608	-.53634	.08144	134.95	6.1628	.08794	.01506
14	S98T002608_D	-.47361	.08296	121.17	6.0577	.07683	.01326
15	S98T002608_S	1.5403	.08035	132.29	10.649	4.2359	4.2199
16	S98T002608_X	-.54077	.14148	149.02	6.6526	.09388	-.00207
17	S98T002608_AX	191.03	.32477	335.28	200.13	202.87	203.48
18	S98T002609	-.62050	.05874	153.94	4.4256	.12390	.02830
19	S98T002609_D	-.60681	.05692	149.10	4.4718	.09883	.01365
20	ICSA	.00422	-.01760	92.169	.23425	-.00381	.00278
21	ICSAB	.45372	-.01393	91.736	-.03848	-.00406	.94164
22	CCV	4.6484	.00610	4.7222	4.6362	4.8857	4.8083
23	CCB	.00001	.00048	-.00011	.23079	.00097	.00061

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.2571	4.8593	5.2581	5.2993	5.0952	4.9095
2	ICB	-.00417	-.00072	.00013	.02616	.00037	-.00047
3	LLS	.21731	.01981	.10264	.24555	.21265	.04206
4	ICSA	253.80	-.00600	-.01626	200.09	.00004	-.01438
5	ICSAB	252.25	.44104	-.01312	198.19	.00133	.92011
6	PREPSTD TJJA	4.6863	4.3923	4.6192	6.0833	4.8842	4.3699
7	PREPBLKTJA	.02247	-.00073	.00151	Q.93870	-.00056	.00401
8	S98T002607	.05037	.01988	.00420	1310.2	.00636	.09761
9	S98T002607_D	.05683	.01781	.00911	1209.1	.01004	.09907
10	CCV	5.2221	4.9056	5.2887	5.0985	4.9894	4.9121
11	CCB	-.00174	-.00053	.00296	.01720	.00107	.00458
12	S98T002608_L	7.3704	1.9209	-.02802	629.89	.83464	64.900
13	S98T002608	6.9156	1.7601	.00916	603.72	.76172	58.444
14	S98T002608_D	6.3595	1.6380	-.00105	622.94	.67281	54.447
15	S98T002608_S	10.978	5.5114	4.0348	589.55	4.8180	60.642
16	S98T002608_X	7.3607	1.9353	-.00574	637.95	.86271	65.854
17	S98T002608_AX	214.82	197.16	197.07	839.39	199.26	258.99
18	S98T002609	7.0184	1.8992	-.02170	684.64	.84953	60.952
19	S98T002609_D	6.7173	1.8403	-.01678	629.08	.79933	58.658
20	ICSA	246.40	-.00484	-.01399	186.83	.00001	-.00190
21	ICSAB	244.83	.43694	-.01613	186.44	-.00060	.89679
22	CCV	5.0294	4.8400	5.2070	4.8149	4.7416	4.8514
23	CCB	-.00121	.00018	.00106	-.00897	-.00008	.00074

Analysis Report

Averages

Tue 01-19-99 01:19:19 PM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.3106	4.9407	5.0325	5.3138	4.8659	5.4354
2	ICB	-.00418	-.00318	.00620	.00529	-.01354	-.00929
3	LLS	.40651	.21854	.20878	.10881	.10529	Q.10529
4	ICSA	.01563	.01154	-.03237	.00894	.03936	-.00629
5	ICSAB	.01588	.97174	-.02796	.00712	.02401	-.00783
6	PREPSTD TJ A	4.5338	4.2647	4.2215	4.4258	4.4278	7.8746
7	PREPBLKTJ A	.04680	.00529	Q.07913	.00578	.01251	Q1.0957
8	S98T002607	5.6862	.06693	3.0372	.02122	.09646	1.4196
9	S98T002607_D	5.5839	.04623	2.4800	.03013	.11900	.81751
10	CCV	5.3524	4.9681	5.0796	5.3259	4.8916	5.4559
11	CCB	.00855	.00142	.00169	-.00004	.00090	-.00262
12	S98T002608_L	92.658	3.4384	5.4037	-.05187	-.17586	6.9127
13	S98T002608	87.151	3.2067	5.0899	-.02768	-.13790	6.5186
14	S98T002608_D	79.020	3.0176	4.5561	-.02865	-.06480	6.0356
15	S98T002608_S	85.897	7.0559	11.893	3.9590	3.9694	13.227
16	S98T002608_X	92.739	3.4937	5.3571	-.04570	.03455	7.0702
17	S98T002608_AX	292.03	200.73	199.90	200.53	194.69	215.90
18	S98T002609	103.86	4.1280	16.020	-.03626	-.08470	12.028
19	S98T002609_D	120.86	3.9632	4.2078	-.02248	-.07043	8.9998
20	ICSA	.02668	.02080	-.02923	.00222	.06694	.06021
21	ICSAB	.04557	.98232	-.02683	.00502	.05776	.04114
22	CCV	5.1492	4.9654	5.0462	5.2515	4.8011	5.4356
23	CCB	.01734	-.00162	-.00279	.01077	.03506	.03818

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	5.0662	5.0626	.17596	5.4052	4.8409	10.220
2	ICB	-.02008	-.00021	.00247	.00119	-.00623	-.06928
3	LLS	.18908	.02081	.01345	.02154	.42973	.44176
4	ICSA	-.03929	.00217	-.00967	.00113	.01666	-.10362
5	ICSAB	-.03530	.00212	-.00753	.00162	.02803	-.10125
6	PREPSTD TJ A	4.8653	4.8454	.14380	4.8135	4.3359	9.5433
7	PREPBLKTJ A	-.04018	-.00025	.00996	.00075	-.02039	-.14007
8	S98T002607	-.08368	.06928	.03475	.01039	.01825	-.03793
9	S98T002607_D	-.07960	.06161	.04653	.00668	.05252	-.05073
10	CCV	4.9781	4.9805	.17507	5.3784	4.8099	10.027
11	CCB	-.01373	-.00008	.00262	.00072	.01638	-.04681
12	S98T002608_L	-.12744	113.42	2.5116	.67004	-.20308	191.61
13	S98T002608	-.03829	114.37	2.1738	.63404	-.03796	185.11
14	S98T002608_D	-.01911	99.235	2.0049	.54467	-.05203	161.47
15	S98T002608_S	4.1238	119.16	2.2571	4.7404	3.8762	176.88
16	S98T002608_X	-.22514	113.61	2.5174	.67987	.08007	190.60
17	S98T002608_AX	198.39	316.72	9.3828	200.53	195.69	578.01
18	S98T002609	.00472	123.67	2.6540	.39783	.03576	213.82
19	S98T002609_D	.02246	118.79	2.6103	.36713	.00462	199.86
20	ICSA	-.01404	.00180	.00090	.00217	.02509	-.13125
21	ICSAB	-.01727	.00177	-.00894	.00165	.02008	-.13406
22	CCV	4.7359	4.8006	.17633	5.2308	4.7098	9.4693
23	CCB	.00273	.00012	.00085	.00094	-.00373	.01701

Analysis Report

Averages

Tue 01-19-99 01:19:19 PM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0268	.00639	4.6933	5.1673
2	ICB	-.00449	-.00106	.00005	-.00301
3	LLS	.09732	-.00097	.02015	.01627
4	ICSA	-.00547	.00615	-.00392	-.00532
5	ICSAB	.46420	.00638	.92336	-.00498
6	PREPSTD TJJA	4.5430	.00464	3.9381	4.5862
7	PREPBLKTJA	-.00883	-.00236	.00521	-.00446
8	S98T002607	-.00897	-.00468	.02013	-.00767
9	S98T002607_D	-.00454	-.00373	.00798	-.00724
10	CCV	5.0368	.00702	4.7555	5.1343
11	CCB	-.00324	-.00076	-.00045	-.00176
12	S98T002608_L	.01206	.06039	.67815	.02239
13	S98T002608	.01825	.05954	.61550	.04865
14	S98T002608_D	.02170	.05517	.56930	.04470
15	S98T002608_S	3.9366	.06649	4.2753	1.1992
16	S98T002608_X	-.01786	.05376	.68641	.01300
17	S98T002608_AX	196.59	.34984	183.39	199.50
18	S98T002609	.05996	.07038	.71492	.06428
19	S98T002609_D	.06407	.06973	.65595	-.03200
20	ICSA	.00109	.00749	-.00359	-.00225
21	ICSAB	.45995	.00750	.92797	-.00158
22	CCV	4.9463	.00780	4.7856	4.9921
23	CCB	.00144	.00030	.00057	.00088

JK LLD

01-19-99

LABCORE Data Entry Template for Worklist# 27739

Analyst: JK Seab Instrument: ICP01-2.3K 1-4-99Book# 77848BMethod: LA-505-151/161 Rev/Mod C-3Ensure dose rate at 30cm
is ≤ 50 mrem/hr prior to
performing this analysisWorklist Comment: ICP BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@ICP-QC	QC		
2	ICB		@ICP-QC	QC		
3	LLS		@ICP-QC	QC		
4	ICSA		@ICP-QC	QC		
5	ICSAB		@ICP-QC	QC		
6	PREPSTDITJA		@ICP-A01	SOLID		
7	PREPBKLTJA		@ICP-A01	SOLID		
8	SAMPLE	S98T002847 0 A	@ICP-A01	SOLID	98000437	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
9	DUP	S98T002847 0 A	@ICP-A01	SOLID		
10	SAMPLE	S98T002926 0 A	@ICP-A01	SOLID	98000490	BY-105 (2)
Analytes Requested: AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01						
11	DUP	S98T002926 0 A	@ICP-A01	SOLID		
12	CCV		@ICP-QC	QC		

Data Entry Comments:

Validated by:
Sal M. Pang
1/19/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27739

S Type	Sample#	R	A	Test	Matrix	Group#	Project
13 CCB				@ICP-QC	QC		
14 SERDIL	S98T002927	0	A	@ICP-A01	SOLID		
15 SAMPLE	S98T002927	0	A	@ICP-A01	SOLID	98000490	BY-105 (2)
Analytes Requested:				AG-A-01 , AL-A-01 , AS-A-01 , B-A-01 , BA-A-01 , BE-A-01 , BI-A-01 , CA-A-01 , CD-A-01 , CE-A-01 , CO-A-01 , CR-A-01 , CU-A-01 , FE-A-01 , K-A-01 , LA-A-01 , LI-A-01 , MG-A-01 , MN-A-01 , MO-A-01 , NA-A-01 , ND-A-01 , NI-A-01 , P-A-01 , PB-A-01 , S-A-01 , SB-A-01 , SE-A-01 , SI-A-01 , SM-A-01 , SR-A-01 , TI-A-01 , TL-A-01 , U-A-01 , V-A-01 , ZN-A-01 , ZR-A-01			
16 DUP	S98T002927	0	A	@ICP-A01	SOLID		
17 SPK-PREDIG	S98T002927	0	A	@ICP-A01	SOLID		
18 ICSA	Sample OK (10 ppm) SEC 1-4-99			@ICP-QC	QC		
19 ICSAB				@ICP-QC	QC		
20 CCV				@ICP-QC	QC		
21 CCB				@ICP-QC	QC		

Final page for worklist # 27739

JK Dd
01-19-99

Analyst Signature	Date
Prep 54 TJG	Direct, 3F 1
Prep 61K TJG	Direct, 1
598T001847	3-6, 3F 3
598T001847-2	3-6, 3
598T001847-4	2-8, 3F 5
598T001847-5	2-8, 5

Analyst Signature	Date
598T001847-6	2-8-2-8, 3F 25
598T001847	2-8, 5
598T001847-2	2-8, 5
598T001847-5	2-8, 5
598T001847-8	2-8-1-9, 50
598T001847-4	2-8-1-9, 50

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1999

$$C_2 = 0.92635 \frac{\mu\text{g}}{\text{ml}} \times \frac{1}{5.1040 \frac{\text{g}}{\text{L}} \times \frac{1\text{L}}{1000 \text{ml}}} = 1.81 \pm 2 \frac{\mu\text{g}}{\text{g}}$$

Pre-dig Spk: $C_2 = \frac{5.9096 - 0.92635}{5} \times 100 = 99.7\%$

Post Spk:

$$\text{Al} = \frac{\left(\frac{513.23}{50}\right) - \left(\frac{25.483}{50}\right)}{10} \times 100 = 97.5\%$$

$$\text{K} = \frac{\left(\frac{485.47}{50}\right) - \left(\frac{3.4222}{50}\right)}{10} \times 100 = 96.4\%$$

$$\text{Na} = \frac{\left(\frac{1826.1}{50}\right) - \left(\frac{1336.3}{50}\right)}{10} \times 100 = 98.0\%$$

$$\text{P} = \frac{\left(\frac{306.96}{50}\right) - \left(\frac{7.0736}{50}\right)}{10} \times 100 = 100.0\%$$

$$\text{Si} = \frac{\left(\frac{500.28}{50}\right) - \left(\frac{2.6453}{50}\right)}{10} \times 100 = 99.5\%$$

Analysis Report

Summary

Tue 01-19-99 10:17:25 AM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990119A	ICP2	01/19/99	08:25	DKS	Q	CONC
2	ICB	990119A	ICP2	01/19/99	08:28	DKS	Q	CONC
3	LLS	990119A	ICP2	01/19/99	08:31	DKS	Q	CONC
4	ICSA	990119A	ICP2	01/19/99	08:35	DKS	Q	CONC
5	ICSAB	990119A	ICP2	01/19/99	08:37	DKS	Q	CONC
6	PREPSTDTJA	990119A	ICP2	01/19/99	08:41	DKS	Q	CONC
7	PREPBLKTJA	990119A	ICP2	01/19/99	08:44	DKS	Q	CONC
8	S98T002847	990119A	ICP2	01/19/99	08:47	DKS	S	CONC
9	S98T002847_D	990119A	ICP2	01/19/99	08:50	DKS	S	CONC
10	S98T002926	990119A	ICP2	01/19/99	08:54	DKS	S	CONC
11	S98T002926_D	990119A	ICP2	01/19/99	08:57	DKS	S	CONC
12	CCV	990119A	ICP2	01/19/99	09:20	DKS	Q	CONC
13	CCB	990119A	ICP2	01/19/99	09:23	DKS	Q	CONC
14	S98T002927_L	990119A	ICP2	01/19/99	09:29	DKS	S	CONC
15	S98T002927	990119A	ICP2	01/19/99	09:33	DKS	S	CONC
16	S98T002927_D	990119A	ICP2	01/19/99	09:36	DKS	S	CONC
17	S98T002927_S	990119A	ICP2	01/19/99	09:40	DKS	S	CONC
18	S98T002927_X	990119A	ICP2	01/19/99	09:45	DKS	S	CONC
19	S98T002927_AX	990119A	ICP2	01/19/99	09:48	DKS	S	CONC
20	ICSA	990119A	ICP2	01/19/99	10:04	DKS	Q	CONC
21	ICSAB	990119A	ICP2	01/19/99	10:07	DKS	Q	CONC
22	CCV	990119A	ICP2	01/19/99	10:10	DKS	Q	CONC
23	CCB	990119A	ICP2	01/19/99	10:14	DKS	Q	CONC

JK RD

01-19-99

Work/4 # 27737

B4-105

598T002847

598T002926

598T002927

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2001 TO 2005

2001

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9085	4.8464	5.0150	5.2497	4.8104	5.0482
2	ICB	-.00055	-.00198	-.00642	.00629	-.00006	-.00013
3	LLS	Q.01417	.10445	.19832	.10404	.09832	.01040
4	ICSA	.00080	240.78	-.03132	-.00173	.00020	.00014
5	ICSAB	.93957	241.46	-.01206	-.00252	.45696	.47302
6	PREPSTDTJA	.86248	4.7384	4.5897	5.1970	4.7442	5.1570
7	PREPBLKTJA	.00001	Q.09702	-.00607	Q.52870	.00054	.00000
8	S98T002847	.08935	87.865	-.09921	.45356	.83764	.00159
9	S98T002847_D	.09695	115.26	-.08883	.44645	1.0958	.00146
10	S98T002926	.07937	206.76	-.05471	.65849	.00919	.00021
11	S98T002926_D	.07508	174.90	-.04527	.52783	.00791	.00070
12	CCV	4.9382	4.9202	5.0274	5.3242	4.8797	5.1401
13	CCB	-.00071	-.00386	-.00292	.00344	-.00001	.00000
14	S98T002927_L 8-19-99	.10610	25.399	.07844	.52911	.00570	-.00233
15	S98T002927	.08476	24.080	-.01626	.50456	.00498	.00068
16	S98T002927_D	.09001	23.816	-.02306	.83186	.00553	.00021
17	S98T002927_S	1.0427	27.083	5.1151	5.6821	4.6519	5.0698
18	S98T002927_X	.11589	25.483	-.25813	.36943	.01042	-.00466
19	S98T002927_AX	424.83	513.23	502.03	497.37	498.59	500.14
20	ICSA	.00503	241.81	-.00773	-.00612	.00030	.00018
21	ICSAB	.94468	243.46	.00849	-.00880	.45934	.47792
22	CCV	4.9082	4.8914	5.0137	5.3125	4.8798	5.1211
23	CCB	.00093	.00173	.00341	.00441	-.00002	.00032

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	4.9571	5.1088	5.0485	4.9752	4.9913	4.9354
2	ICB	.01463	-.00726	-.00073	-.00821	-.00174	-.00189
3	LLS	.17157	.22374	.01197	.20701	.04107	.02097
4	ICSA	-.00970	252.84	.00596	-.00502	-.00138	-.00384
5	ICSAB	-.02406	253.73	.93745	.00122	.45249	.45847
6	PREPSTDTJA	4.5718	4.8563	4.5944	4.9068	4.6400	4.6493
7	PREPBLKTJA	-.00475	Q.19164	-.00029	-.00369	-.00084	.00074
8	S98T002847	6.1398	48.619	.01402	-.02816	.04235	4.8574
9	S98T002847_D	7.1191	166.98	.01771	-.01008	.04802	5.4671
10	S98T002926	.03683	.81768	.01923	-.01261	.00121	3.2856
11	S98T002926_D	-.04275	.79812	.01761	-.01237	.00130	2.8009
12	CCV	5.0672	5.1042	5.0815	5.0622	5.0195	4.9655
13	CCB	.01463	-.00628	.00022	-.00072	-.000163	-.00090
14	S98T002927_L 8-19-99	-.00155	.39226	.00678	.04100	-.02399	.97090
15	S98T002927	-.04124	.58067	.00509	.00870	.00530	.92635
16	S98T002927_D	.00564	.69162	.00882	.00551	-.01019	.98086
17	S98T002927_S	5.1093	5.7073	5.1142	4.9419	5.0144	5.9096
18	S98T002927_X	.35149	.23392	.03049	.21833	-.02028	1.0411
19	S98T002927_AX	506.35	501.98	496.68	503.41	496.93	499.09
20	ICSA	-.00676	252.50	.00513	-.00039	-.00018	-.00201
21	ICSAB	-.00166	253.44	.93780	-.00369	.45393	.46302
22	CCV	4.9677	5.0729	5.0451	5.0611	4.9794	4.9353
23	CCB	-.00774	-.00903	-.00065	-.00477	-.00016	-.00028

Analysis Report

Averages

Tue 01-19-99 10:17:25 AM

page 3

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7516	.00638	4.7680	4.6601	4.9904	4.9706
2	ICB	-.00086	-.00001	-.00081	-.48509	-.00082	-.00010
3	LLS	.01780	.00035	.09755	Q.29765	.10176	.02088
4	ICSA	.00277	-.00461	91.679	-.42331	-.00440	.00201
5	ICSAB	.46286	-.00308	91.884	-.22825	-.00395	.96830
6	PREPSTDJTJA	4.4997	.00489	4.5294	4.2776	4.9138	4.9463
7	PREPBLKTJA	.00278	-.00152	.02540	-.47661	-.00063	-.00110
8	S98T002847	-.20035	.04477	59.585	5.4840	.03223	.00625
9	S98T002847_D	-.22500	.13632	69.512	7.2902	.04086	.00950
10	S98T002926	.03306	.00306	1.4164	3.2973	-.00056	.00197
11	S98T002926_D	.02476	.00743	1.2405	2.3964	-.00193	.00649
12	CCV	4.8606	.00453	4.8098	5.1886	5.0836	5.0743
13	CCB	.00041	-.00084	-.00094	-.10854	-.00070	-.00123
14	S98T002927_L 建 1-19-99	.01578	-.01710	.50203	.82085	-.00197	-.01796
15	S98T002927	.00906	.00066	.46656	1.7123	-.00110	.00151
16	S98T002927_D	.00619	.00334	.61442	2.0390	.00047	.00254
17	S98T002927_S	4.5700	.00867	5.3932	6.4352	4.8889	4.7349
18	S98T002927_X	.05440	.01022	.56180	3.4222	.04575	-.01032
19	S98T002927_AX	473.32	.47739	478.77	485.47	501.79	493.03
20	ICSA	.00414	-.00942	92.062	-.16389	-.00456	.00195
21	ICSAB	.46825	-.00655	92.482	-.15369	-.00485	.97255
22	CCV	4.8577	.00474	4.7695	4.8411	5.0786	5.0427
23	CCB	.00067	-.00186	-.00068	-.26517	-.00079	-.00163

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1530	4.8876	5.2454	5.0367	4.8635	4.8941
2	ICB	-.00644	.00003	-.00051	-.00033	-.00399	.00235
3	LLS	.21273	.02064	.10198	.20288	.20156	.04908
4	ICSA	246.03	-.00474	-.01406	189.43	-.00286	-.00695
5	ICSAB	246.03	.43987	-.01514	189.77	-.00511	.92330
6	PREPSTDJTJA	4.7703	4.5650	4.7841	5.7540	4.8680	4.5200
7	PREPBLKTJA	.03238	.00041	.00084	Q.85533	-.00114	.00840
8	S98T002847	2.8666	.83910	.03617	867.39	.31033	25.331
9	S98T002847_D	4.5126	.94477	.02973	901.64	.35806	30.205
10	S98T002926	.12649	.06846	.03181	1111.1	.01158	.44710
11	S98T002926_D	.10740	.05594	.02743	1054.5	.00525	.42733
12	CCV	5.2170	4.9068	5.2935	5.0949	4.9691	4.9439
13	CCB	-.00930	-.00040	.00068	.00431	-.00188	.00284
14	S98T002927_L 建 1-19-99	.04246	.01333	.01745	1355.6	.02276	.30653
15	S98T002927	.04433	.02321	.01253	1244.3	.00978	.34612
16	S98T002927_D	.05962	.02708	.01260	1278.6	.00686	.40741
17	S98T002927_S	5.1845	4.8370	5.1245	1278.8	4.7842	5.2551
18	S98T002927_X	.07156	.02344	.03552	1336.3	.11101	.43228
19	S98T002927_AX	516.78	494.40	498.95	1826.1	490.34	495.09
20	ICSA	247.89	-.00528	-.01468	190.19	.00066	.00151
21	ICSAB	248.73	.43960	-.01576	192.34	-.00289	.91035
22	CCV	5.1553	4.8661	5.2525	5.0552	4.9616	4.9045
23	CCB	-.00874	-.00068	.00031	.02581	-.00098	-.00133

Analysis Report

Averages

Tue 01-19-99 10:17:25 AM

page 4

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.2477	4.9866	5.0799	5.2626	4.8043	5.3833
2	ICB	.01038	-.01041	.00749	-.00448	.01136	.00207
3	LLS	.39433	.20421	.20708	.09618	.20484	.11910
4	ICSA	.02982	.00476	-.01618	.00807	.05019	.00219
5	ICSAB	.02789	.98625	-.01065	.00542	.03617	-.00249
6	PREPSTD TJ A	4.6597	4.4380	4.3777	4.5575	4.5454	7.7052
7	PREPBLKTJA	.04224	-.00085	Q.09398	-.00257	.03411	Q.69296
8	S98T002847	35.443	1.5485	5.8372	-.05343	.02993	11.113
9	S98T002847_D	71.652	1.8020	18.022	-.04457	-.04191	33.961
10	S98T002926	9.1830	.15188	6.0294	-.01157	.25763	2.8000
11	S98T002926_D	27.212	.06213	6.4921	-.05925	.31381	2.2700
12	CCV	5.3277	4.9894	5.1230	5.3258	4.8709	5.4674
13	CCB	.00176	-.00472	.01009	-.00371	.00204	-.00311
14	S98T002927_L 1-19-99	6.9247	.01641	15.307	.14155	-.06231	2.0272
15	S98T002927	6.6602	.02832	14.844	.00714	.06009	1.6451
16	S98T002927_D	6.5909	.04807	16.960	.03533	.07354	2.2663
17	S98T002927_S	13.200	5.1257	20.060	5.1212	4.9673	10.243
18	S98T002927_X	7.0736	.45257	15.766	.07948	-.18128	2.6453
19	S98T002927_AX	506.96	498.60	502.21	503.54	487.54	500.28
20	ICSA	.03109	.01203	-.02676	.01130	.03072	.00012
21	ICSAB	.03499	.98173	-.02037	.00819	.03280	-.00492
22	CCV	5.2696	4.9430	5.0707	5.3129	4.8347	5.4309
23	CCB	.00131	-.00471	.00537	.00358	-.01200	-.00816

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8431	4.8591	.18094	5.2834	4.7439	9.7376
2	ICB	.00045	-.00004	-.00429	-.00070	-.01687	-.00070
3	LLS	.19960	.01997	.00770	.02041	.35795	.49201
4	ICSA	-.02047	.00179	-.00637	.00070	-.02355	-.12948
5	ICSAB	-.02337	.00176	-.00011	.00119	-.01066	-.13930
6	PREPSTD TJ A	4.8560	4.8537	.16684	4.8851	4.4212	9.4796
7	PREPBLKTJA	-.01927	.00016	-.00098	.00115	.00037	-.06645
8	S98T002847	-.00381	38.181	1.0911	.24642	-.10590	74.090
9	S98T002847_D	-.00758	44.446	1.2696	.36419	-.03214	86.525
10	S98T002926	.01157	.08874	.04735	.03919	-.11478	.06031
11	S98T002926_D	.06764	.07144	-.01578	.02860	-.08885	.65019
12	CCV	4.9477	4.9672	.16980	5.3637	4.8210	9.9620
13	CCB	-.00838	-.00008	-.00193	-.00023	.01230	-.03585
14	S98T002927_L 1-19-99	-.20540	.03055	.07245	.01188	.18479	-.25376
15	S98T002927	-.00080	.03092	-.00560	.01273	.02799	.39603
16	S98T002927_D	.02901	.03593	-.00101	.01863	.03679	.50770
17	S98T002927_S	4.7613	4.8842	.16078	5.0621	4.7825	9.7047
18	S98T002927_X	.08209	.03481	.02574	.07044	1.5017	.92225
19	S98T002927_AX	486.60	498.59	16.738	500.08	488.68	956.31
20	ICSA	-.02559	.00181	-.00248	.00165	-.02028	-.09673
21	ICSAB	-.02739	.00196	-.00914	.00092	.03040	-.10608
22	CCV	4.9368	4.9629	.17099	5.3505	4.8039	9.9453
23	CCB	-.01929	-.00025	.00200	.00072	-.01290	-.06801

Analysis Report

Averages

Tue 01-19-99 10:17:25 AM

page 5

#	Sample Name	V	Y	Zn	Zr
1	ICV	4.9852	.00745	4.7727	5.0590
2	ICB	-.00053	-.00023	.00003	.00032
3	LLS	.10113	.00026	.01925	.01898
4	ICSA	-.00092	.00695	-.00469	-.00301
5	ICSAB	.46039	.00695	.92096	-.00391
6	PREPSTD TJ A	4.6627	.00576	4.1406	4.6910
7	PREPBLK TJ A	-.00372	-.00153	.00383	-.00113
8	S98T002847	.02095	.02751	.26287	.02125
9	S98T002847_D	.03843	.03797	.30670	.02602
10	S98T002926	.02123	.00052	.02987	.03902
11	S98T002926_D	.07250	.00364	.02592	.02763
12	CCV	5.0331	.00712	4.7905	5.1254
13	CCB	-.00288	-.00045	-.00041	-.00110
14	S98T002927_L <i>1-18-99</i>	-.04140	-.00772	.00534	-.02930
15	S98T002927	.01670	.00139	.02786	.00812
16	S98T002927_D	.02014	.00291	.02738	.01413
17	S98T002927_S	4.9218	.01037	4.8591	4.8112
18	S98T002927_X	.04419	.01136	.01874	.01319
19	S98T002927_AX	495.38	.74148	468.62	496.96
20	ICSA	-.00132	.00701	-.00438	-.00325
21	ICSAB	.46233	.00686	.92520	-.00321
22	CCV	5.0079	.00705	4.7517	5.1095
23	CCB	-.00486	-.00106	.00018	-.00265

JK LdB
01-19-99

2005

LABCORE Data Entry Template for Worklist# 27861

Analyst: MSG Instrument: ICP01 1-13-99

Book# 77848B

Method: LA-505-151/161 Rev/Mod C-3

**Ensure dose rate at 30cm
is $\leq$ 50 mrem/hr prior to
performing this analysis**

Worklist Comment: 1-13-99
ICP BY-105 (FUSION)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@ICP-QC	QC		
2 ICB			@ICP-QC	QC		
3 LLS			@ICP-QC	QC		
4 ICSA			@ICP-QC	QC		
5 ICSAB			@ICP-QC	QC		
6 PREPBLKTJA			@ICP-F01	SOLID		
7 SERDIL	S98T003393	0 F	@ICP-F01	SOLID		
8 SAMPLE	S98T003393	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 , MN-F-01 , MO-F-01 , NA-F-01 , ND-F-01 , NI-F-01 , P-F-01 , PB-F-01 , S-F-01 , SB-F-01 , SE-F-01 , SI-F-01 , SM-F-01 , SR-F-01 , TI-F-01 , TL-F-01 , U-F-01 , V-F-01 , ZN-F-01 , ZR-F-01						
9 DUP	S98T003393	0 F	@ICP-F01	SOLID		
10 SPK <i>(1pph)</i>	S98T003393	0 F	@ICP-F01	SOLID		
11 CCV <i>(10 ppm)</i>			@ICP-QC	QC		
12 CCB			@ICP-QC	QC		
13 SERDIL	S98T003398	0 F	@ICP-F01	SOLID		
14 SAMPLE	S98T003398	0 F	@ICP-F01	SOLID	98000436	BY-105 (2)
Analytes Requested: AG-F-01 , AL-F-01 , AS-F-01 , B-F-01 , BA-F-01 , BE-F-01 , BI-F-01 , CA-F-01 , CD-F-01 , CE-F-01 , CO-F-01 , CR-F-01 , CU-F-01 , FE-F-01 , LA-F-01 , LI-F-01 , MG-F-01 ,						

Data Entry Comments:

uploaded 1-8-99

John Wauell

27861-1-18-99

Validated by:

Sail M. Parag

1/18/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

01/12/99 13:40
A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27861

S Type	Sample#	R A	Test	Matrix	Group#	Project		
			MN-F-01	MO-F-01	NA-F-01	ND-F-01	NI-F-01	P-F-01
			PB-F-01	S-F-01	SB-F-01	SE-F-01	SI-F-01	SM-F-01
			SR-F-01	TI-F-01	TL-F-01	U-F-01	V-F-01	ZN-F-01
			ZR-F-01					
15 DUP	S98T003398	0 F	@ICP-F01	SOLID				
16 SPK (1ppm)	S98T003398	0 F	@ICP-F01	SOLID				
17 SERDIL	S98T003388	0 F	@ICP-F01	SOLID				
18 SAMPLE	S98T003388	0 F	@ICP-F01	SOLID	98000490	BY-105 (2)		
Analytes Requested:								
			AG-F-01	AL-F-01	AS-F-01	B-F-01	BA-F-01	
			BE-F-01	BI-F-01	CA-F-01	CD-F-01	CE-F-01	
			CR-F-01	CU-F-01	FE-F-01	LA-F-01	LI-F-01	
			MN-F-01	MO-F-01	NA-F-01	ND-F-01	NI-F-01	
			PB-F-01	S-F-01	SB-F-01	SE-F-01	SI-F-01	
			SR-F-01	TI-F-01	TL-F-01	U-F-01	V-F-01	
			ZR-F-01					
19 DUP	S98T003388	0 F	@ICP-F01	SOLID				
20 SPK (1ppm)	S98T003388	0 F	@ICP-F01	SOLID				
21 ICSA			@ICP-QC	QC				
22 ICSAB			@ICP-QC	QC				
23 CCV			@ICP-QC	QC				
24 CCB			@ICP-QC	QC				

Final page for worklist # 27861

01-18-99

Analyst Signature	Date	Analyst Signature	Date
Prep 6/17/99	15-10-18	5987003398-L	15-10-18
5987003393-L	15-10-18	5987003398	15-10
5987003393	15-10	5987003398-D	15-10
5987003393-D	15-10	5987003398-G	15-10
5987003393-X	15-10	5987003398-X	15-10-19
5987003393-GX	15-10-19	5987003398-GX	15-10-19
5987003393-L	15-10-18	5987003388-L	15-10-18
5987003393	15-10	5987003388	15-10
5987003393-D	15-10	5987003388-D	15-10
5987003393-X	15-10	5987003388-G	15-10
5987003393-GX	15-10-19	5987003388-X	15-10-19
		5987003388-GX	15-10-19

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

$$Al = 16.830 \frac{\mu g}{ml} \times \frac{1}{2.0140 \frac{g}{L} \times \frac{1L}{1000 ml}} = 8.36 \mu g + 3 \frac{\mu g}{g}$$

$$Spike: Al = \frac{\left(\frac{56.399}{41}\right) - \left(\frac{16.830}{41}\right)}{1} \times 100 = 97.0\%$$

$$Post Spike: Na = \frac{\left(\frac{4574.2}{410}\right) - \left(\frac{519.04}{410}\right)}{10} \times 100 = 98.9\%$$

S98T003398

$$Al = 49.817 \frac{\mu g}{ml} \times \frac{1}{1.9860 \frac{g}{L} \times \frac{1L}{1000 ml}} = 2.51 \mu g + 4 \frac{\mu g}{g}$$

$$Spike: Al = \frac{\left(\frac{3995.6}{410}\right) - \left(\frac{54.227}{41}\right) - \left(\frac{49.817}{41}\right)}{1} \times 100 =$$

sp 1/18/99

$$Al = \frac{\left(\frac{88.212}{41}\right) - \left(\frac{49.817}{41}\right)}{1} \times 100 = 93.6\%$$

S98T003388

$$Al = 32.587 \frac{\mu g}{ml} \times \frac{1}{2.0220 \frac{g}{L} \times \frac{1L}{1000 ml}} = 1.61 \mu g + 4 \frac{\mu g}{g}$$

$$Spike: \frac{\left(\frac{70.843}{41}\right) - \left(\frac{32.587}{41}\right)}{1} \times 100 = 93.3\%$$

Post Spike:

$$Na = \frac{\left(\frac{4370.7}{410}\right) - \left(\frac{478.51}{410}\right)}{10} \times 100 = 94.9\%$$

Analysis Report

Summary

Mon 01-18-99 12:21:49 PM

page 1

#	Sample Name	File	Method	Date	Time	OpID	Type	Mode
1	ICV	990118A	ICP2	01/18/99	08:38	DKS	Q	CONC
2	ICB	990118A	ICP2	01/18/99	08:59	DKS	Q	CONC
3	LLS	990118A	ICP2	01/18/99	09:02	DKS	Q	CONC
4	ICSA	990118A	ICP2	01/18/99	09:04	DKS	Q	CONC
5	ICSAB	990118A	ICP2	01/18/99	09:08	DKS	Q	CONC
6	PREPBLKTJA	990118A	ICP2	01/18/99	09:12	DKS	Q	CONC
7	S98T003393_L	990118A	ICP2	01/18/99	09:15	DKS	S	CONC
8	S98T003393	990118A	ICP2	01/18/99	09:19	DKS	S	CONC
9	S98T003393_D	990118A	ICP2	01/18/99	09:26	DKS	S	CONC
10	S98T003393_A	990118A	ICP2	01/18/99	09:30	DKS	S	CONC
11	S98T003393_X	990118A	ICP2	01/18/99	09:34	DKS	S	CONC
12	S98T003393_AX	990118A	ICP2	01/18/99	09:38	DKS	S	CONC
13	CCV	990118A	ICP2	01/18/99	09:52	DKS	Q	CONC
14	CCB	990118A	ICP2	01/18/99	09:55	DKS	Q	CONC
15	S98T003398_L	990118A	ICP2	01/18/99	09:59	DKS	S	CONC
16	S98T003398	990118A	ICP2	01/18/99	10:02	DKS	S	CONC
17	S98T003398_D	990118A	ICP2	01/18/99	10:06	DKS	S	CONC
18	S98T003398_A	990118A	ICP2	01/18/99	10:09	DKS	S	CONC
19	S98T003398_X	990118A	ICP2	01/18/99	10:15	DKS	S	CONC
20	S98T003398_AX	990118A	ICP2	01/18/99	10:18	DKS	S	CONC
21	S98T003388_L	990118A	ICP2	01/18/99	10:36	DKS	S	CONC
22	S98T003388	990118A	ICP2	01/18/99	10:39	DKS	S	CONC
23	S98T003388_D	990118A	ICP2	01/18/99	10:42	DKS	S	CONC
24	S98T003388_A	990118A	ICP2	01/18/99	10:46	DKS	S	CONC
25	S98T003388_X	990118A	ICP2	01/18/99	10:50	DKS	S	CONC
26	S98T003388_AX	990118A	ICP2	01/18/99	10:53	DKS	S	CONC
27	ICSA	990118A	ICP2	01/18/99	12:08	DKS	Q	CONC
28	ICSAB	990118A	ICP2	01/18/99	12:11	DKS	Q	CONC
29	CCV	990118A	ICP2	01/18/99	12:14	DKS	Q	CONC
30	CCB	990118A	ICP2	01/18/99	12:18	DKS	Q	CONC

JK [Signature]

01-18-99

Worklist # 27861

84-105

5887003393

5887003398

5887003388

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES ~~2009~~ TO 2014

Analysis Report

Averages

Mon 01-18-99 12:21:49 PM

page 2

#	Sample Name	Ag	Al	As	B	Ba	Be
1	ICV	4.9345	4.8670	5.0331	5.3877	4.8239	5.0665
2	ICB	.00087	.00260	.00586	.00449	.00007	.00009
3	LLS	Q.01486	.11001	.20288	.10638	.09946	.01069
4	ICSA	.00353	243.74	-.02189	-.00590	.00038	.00014
5	ICSAB	.95027	241.31	.02092	-.00645	.45478	.47375
6	PREPBLKTJA	.00218	.00893	-.00465	.00647	.00014	-.00004
7	S98T003393_L	.36108	18.271	.77764	.30631	.02166	-.00941
8	S98T003393	.10184	16.830	.19646	-.04137	.01041	.00765
9	S98T003393_D	.09064	17.801	-.14218	.18396	.00584	.00583
10	S98T003393_A	37.776	56.599	41.227	42.853	37.863	40.510
11	S98T003393_X	.68793	21.337	3.5737	2.6587	.07343	.03757
12	S98T003393_AX	3777.6	4013.5	4154.1	4158.8	4098.8	4085.6
13	CCV	4.9189	4.8380	5.0315	5.3532	4.8214	5.0620
14	CCB	.00341	.01051	.00354	.00698	.00010	.00027
15	S98T003398_L	.30293	50.872	.78638	1.5743	1.0263	.04268
16	S98T003398	.11583	49.817	.45381	.18141	.98601	.00535
17	S98T003398_D	.16514	48.976	.29440	.32376	.95581	.00358
18	S98T003398_A	37.678	88.212	40.985	42.399	38.224	40.025
19	S98T003398_X	1.1035	54.227	-.09586	.86092	1.0758	.07068
20	S98T003398_AX	3682.0	3995.6	4131.4	4094.7	4028.8	4031.9
21	S98T003388_L	.15613	32.852	-.30273	.20384	.00816	.03794
22	S98T003388	.14838	32.587	.12588	.22491	.01340	.00203
23	S98T003388_D	.06653	30.263	.23895	.06160	.01141	.00016
24	S98T003388_A	37.043	70.843	41.210	42.381	36.902	39.626
25	S98T003388_X	.98670	33.653	-.58097	2.2568	.03985	-.00006
26	S98T003388_AX	3403.4	3945.3	4111.0	4050.8	3977.1	4002.2
27	ICSA	.00272	239.65	.00215	-.00979	.00029	.00033
28	ICSAB	.94903	240.00	.04809	-.01199	.45184	.47242
29	CCV	4.9567	4.8222	5.0318	5.3302	4.7764	5.0520
30	CCB	.00216	.00629	-.00151	.00599	.00022	.00009

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
1	ICV	5.0212	5.0944	5.0851	5.0181	5.0149	4.9795
2	ICB	-.04115	-.02160	.00184	.00637	.00180	.00238
3	LLS	.18100	.20249	Q.01258	.21599	.04222	.02093
4	ICSA	-.02413	254.08	.00865	.01438	.00327	-.00014
5	ICSAB	-.05666	255.07	.94827	.01443	.46107	.46536
6	PREPBLKTJA	-.02597	-.02133	.00254	.01456	.00134	.00100
7	S98T003393_L	-9.3538	-4.3751	.35896	2.2476	.36730	.65504
8	S98T003393	-1.3039	-.28896	.09392	.66769	.17488	.34414
9	S98T003393_D	-.91272	-.80189	.11149	.37924	.00496	.25559
10	S98T003393_A	38.977	40.691	41.583	40.614	40.695	40.603
11	S98T003393_X	-11.745	-9.6238	1.0081	5.5982	1.9838	1.3278
12	S98T003393_AX	4177.7	4108.4	4106.1	4148.4	4092.2	4105.9
13	CCV	5.0071	5.0681	5.0809	5.0281	5.0100	4.9625
14	CCB	-.00941	-.02196	.00176	.00666	.00180	.00100
15	S98T003398_L	1.5511	49.625	.41281	1.3583	1.6922	1.6922
16	S98T003398	5.8151	54.120	.03301	.13923	.11111	1.4030
17	S98T003398_D	6.0956	56.136	.10978	.73814	.06408	1.2934
18	S98T003398_A	46.992	96.845	41.562	40.107	41.084	41.930
19	S98T003398_X	-9.3226	46.602	1.0042	7.1705	.68611	2.4009

Analysis Report Averages Mon 01-18-99 12:21:49 PM page 3

#	Sample Name	Bi	Ca	Cd	Ce	Co	Cr
20	S98T003398_AX	4192.2	4224.2	4117.6	4068.3	4111.3	4135.9
21	S98T003388_L	-8.2401	-5.5689	.25771	.41135	.56672	.63894
22	S98T003388	-.21453	-.15116	.07827	.58673	.10144	.67855
23	S98T003388_D	-2.4685	-.85836	.05410	.26295	.06229	.64170
24	S98T003388_A	39.477	41.353	41.450	39.487	40.713	41.055
25	S98T003388_X	-12.862	-9.9696	1.1516	4.5376	.12532	.44571
26	S98T003388_AX	4161.3	4187.5	4114.9	4017.0	4109.2	4134.3
27	ICSA	-.00898	252.03	.00881	.01261	.00190	-.00142
28	ICSAB	-.05880	255.15	.95016	.01346	.46024	.46878
29	CCV	4.9852	5.1368	5.1201	5.0031	5.0503	5.0107
30	CCB	-.04269	-.01925	.00247	.01070	.00236	.00182

#	Sample Name	Cu	Eu	Fe	K	La	Li
1	ICV	4.7799	.00675	4.7935	5.1397	5.0168	4.9932
2	ICB	.00073	.00072	-.00166	.46606	.00124	.00080
3	LLS	.01813	.00073	.09665	Q.81901	.10413	.02209
4	ICSA	.00408	-.00664	92.894	.38548	-.00208	.00243
5	ICSAB	.46379	-.00843	92.356	.56114	-.00189	.96778
6	PREPBLKTJA	.00033	.00066	.00473	Q248.02	.00193	.00030
7	S98T003393_L	.14543	.05057	.17248	11445.	.37391	.02072
8	S98T003393	.06814	.04156	.35639	10896.	.08346	.02906
9	S98T003393_D	.00827	.02400	.37947	10819.	.05649	.02496
10	S98T003393_A	38.371	.08018	38.948	11201.	40.141	38.954
11	S98T003393_X	.35181	.25662	.12267	11588.	.66692	.20669
12	S98T003393_AX	3866.6	4.9798	3930.5	15398.	4128.8	4099.8
13	CCV	4.7684	.00668	4.7752	5.1150	5.0190	4.9939
14	CCB	.00060	.00059	-.00106	.40017	.00086	.00050
15	S98T003398_L	-.15904	.18448	65.117	11399.	.20003	.05760
16	S98T003398	-.16080	.09498	65.041	10994.	.06984	.04552
17	S98T003398_D	-.17659	.13544	62.992	10366.	.11503	.06216
18	S98T003398_A	37.372	.17261	103.43	10850.	39.443	37.856
19	S98T003398_X	-.14675	.82150	66.316	11316.	1.0189	.69854
20	S98T003398_AX	3768.7	5.4105	3991.2	14974.	4036.7	3985.4
21	S98T003388_L	.08709	.10959	-.10596	11147.	.07344	.08275
22	S98T003388	.05439	.05933	.38408	10732.	.08037	.03733
23	S98T003388_D	-.00007	.11218	.22282	10385.	.06447	.08690
24	S98T003388_A	36.947	.14367	38.741	10530.	38.911	36.983
25	S98T003388_X	-.13421	.90597	-.45076	10806.	.71569	.82668
26	S98T003388_AX	3714.5	5.2974	3916.8	14448.	3986.4	3885.5
27	ICSA	.00386	-.00852	91.728	.32646	-.00246	.00283
28	ICSAB	.45897	-.01092	92.188	.40867	-.00224	.95757
29	CCV	4.7206	.00754	4.7786	4.9396	4.9882	4.9236
30	CCB	.00026	.00144	-.00138	.30451	.00267	.00110

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
1	ICV	5.1812	4.9010	5.2682	4.9552	4.8837	4.9385
2	ICB	.00915	-.00008	.00180	-.00317	.00520	-.00103
3	LLS	.22237	.02058	.10341	.20673	.20510	.04172
4	ICSA	248.51	-.00462	-.01355	192.48	.00161	-.01385
5	ICSAB	245.32	.44257	-.01608	189.41	.00002	.90919

Analysis Report

Averages

Mon 01-18-99 12:21:49 PM

page 4

#	Sample Name	Mg	Mn	Mo	Na	Nd	Ni
6	PREPBLKTJA	.01564	.00098	.00140	.06667	.00067	Q.07423
7	S98T003393_L	1.9785	-.01372	.36617	517.16	.47769	2.0965
8	S98T003393_	.60560	.04502	.09448	493.82	.17532	1.8447
9	S98T003393_D	.26002	.04260	.02825	507.75	-.05024	1.2102
10	S98T003393_A	41.608	39.116	42.932	546.67	38.907	41.762
11	S98T003393_X	5.7130	-.00427	.58879	519.04	1.8567	3.3980
12	S98T003393_AX	4227.7	4079.2	4091.6	4574.2	4022.3	4094.0
13	CCV	5.1299	4.8924	5.2588	4.9510	4.8868	4.9290
14	CCB	.00976	.00002	.00162	.00492	.00217	-.00167
15	S98T003398_L	3.8640	.89756	-.02639	268.87	.91460	39.249
16	S98T003398_	3.0899	.87987	.08867	259.28	.49422	38.569
17	S98T003398_D	3.4672	.89403	.05838	257.13	.45356	43.409
18	S98T003398_A	43.710	40.021	43.126	294.93	38.671	79.021
19	S98T003398_X	8.2942	1.0593	-.01750	256.93	1.4884	41.725
20	S98T003398_AX	4173.2	4099.3	4100.9	4232.5	3923.0	4151.6
21	S98T003388_L	1.0012	.04616	.22120	508.16	.12760	1.1510
22	S98T003388_	.60733	.07042	.07984	489.24	.20971	1.7236
23	S98T003388_D	.21940	.08595	.05920	472.41	.10127	2.9705
24	S98T003388_A	40.769	39.119	42.824	513.28	37.606	41.841
25	S98T003388_X	3.5576	.29957	.89849	478.51	.62887	2.9239
26	S98T003388_AX	4124.3	4078.7	4095.7	4370.7	3872.2	4110.7
27	ICSA	243.41	-.00423	-.01247	188.88	.00147	-.00800
28	ICSAB	243.17	.44201	-.01321	187.76	.00060	.91950
29	CCV	5.0827	4.9187	5.3086	5.3276	4.8495	4.9726
30	CCB	.01308	.00031	.00163	.00634	.00389	-.00053

#	Sample Name	P	Pb	S	Sb	Se	Si
1	ICV	5.2553	5.0037	5.1164	5.2979	4.8157	5.3720
2	ICB	-.00947	.00898	.00854	.00316	.00489	-.00073
3	LLS	.37926	.22758	.20618	.10192	.21590	.11299
4	ICSA	.03215	.02494	-.03124	.00963	.04970	-.00193
5	ICSAB	.02260	.99375	-.02718	-.00803	.06429	-.00271
6	PREPBLKTJA	.00698	.00295	.01066	.00175	.00925	.00820
7	S98T003393_L	-.31818	3.6075	-.08334	.94434	.72945	.78114
8	S98T003393_	1.3037	.45072	.72247	.00656	.51407	1.4300
9	S98T003393_D	1.1603	.13609	1.1288	-.05557	.65707	1.3734
10	S98T003393_A	43.572	40.929	42.126	43.603	40.369	45.861
11	S98T003393_X	4.5491	5.3673	2.7963	1.0804	2.1059	1.3353
12	S98T003393_AX	4075.2	4126.2	4021.5	4128.5	4013.1	4004.5
13	CCV	5.2252	5.0162	5.0613	5.2868	4.8124	5.3465
14	CCB	.00284	.00549	.01091	.00286	.02687	-.00118
15	S98T003398_L	43.143	4.2388	3.3308	.46117	4.5242	15.373
16	S98T003398_	43.520	2.1215	2.6348	-.06262	.83214	15.078
17	S98T003398_D	41.220	2.0653	2.4522	-.20319	1.4853	15.733
18	S98T003398_A	85.555	43.188	44.984	43.939	39.906	59.558
19	S98T003398_X	45.221	5.4387	2.2328	2.4896	14.076	17.887
20	S98T003398_AX	4102.8	4161.5	4034.0	4121.2	3998.8	4023.9
21	S98T003388_L	4.6061	.98962	2.5922	.00511	3.8397	2.1444
22	S98T003388_	3.4631	.68930	3.6957	.03879	1.3163	2.0737
23	S98T003388_D	3.6092	.12816	3.9988	.02362	1.8079	2.1672
24	S98T003388_A	46.616	41.654	45.617	43.520	40.539	46.315
25	S98T003388_X	11.344	2.6876	.23221	.40114	16.224	6.5666

2012

Analysis Report

Averages

Mon 01-18-99 12:21:49 PM

page 5

#	Sample Name	P	Pb	S	Sb	Se	Si
26	S98T003388_AX	4044.6	4166.2	4019.6	4103.4	3981.9	3995.0
27	ICSA	.03650	.02414	-.01350	-.00174	.07890	-.00353
28	ICSAB	.03810	.99360	-.02669	-.00115	.08260	-.00138
29	CCV	5.2631	5.0525	5.1447	5.3019	4.8643	5.3608
30	CCB	.00482	.01331	.00117	.00406	.03718	-.00046

#	Sample Name	Sm	Sr	Th	Ti	Tl	U
1	ICV	4.8723	4.8938	.20293	5.3100	4.7554	9.8227
2	ICB	.00781	.00000	.00319	.00094	.00412	.03307
3	LLS	.20555	.02036	.00979	.02087	.43493	.50570
4	ICSA	-.01506	.00182	.00275	.00169	.02912	-.14327
5	ICSAB	-.01274	.00179	.00885	.00149	.01424	-.14654
6	PREPBLKTJA	.00280	.00009	.01250	.00054	.02370	-.05100
7	S98T003393_L	.40921	.01797	.89884	.05265	-.61893	-1.4473
8	S98T003393	.39580	.02667	.38299	.04103	-.23375	-1.3171
9	S98T003393_D	.15993	.01613	.42351	-.00668	.46265	-2.0695
10	S98T003393_A	38.670	39.016	2.0065	42.846	37.839	76.087
11	S98T003393_X	2.1974	.03509	2.9875	.39410	7.9980	7.2749
12	S98T003393_AX	4005.8	4085.5	160.77	4088.4	4065.9	7814.6
13	CCV	4.8841	4.8881	.19638	5.3004	4.8071	9.8115
14	CCB	.00335	.00000	.00650	.00095	.01803	.01817
15	S98T003398_L	.98709	47.359	1.9449	.30129	-1.4260	83.772
16	S98T003398	.30647	46.448	1.4112	.28235	-.50952	77.434
17	S98T003398_D	.77552	45.310	1.6204	.24621	.69901	76.052
18	S98T003398_A	38.218	84.287	3.2056	42.494	37.700	153.45
19	S98T003398_X	6.0029	46.889	4.3943	.35594	8.7982	103.14
20	S98T003398_AX	3895.1	4037.2	157.43	4030.8	4023.7	7669.4
21	S98T003388_L	.77579	.02660	.42728	.09778	1.7747	1.5382
22	S98T003388	.52332	.03372	.41157	.06992	.55296	-1.0196
23	S98T003388_D	1.0762	.03708	.07537	.02994	-.17150	1.2865
24	S98T003388_A	37.879	37.802	1.7431	41.771	37.719	75.181
25	S98T003388_X	7.9879	.12149	.75954	-.00134	3.0477	31.137
26	S98T003388_AX	3830.9	3948.7	154.14	4005.4	4003.4	7468.8
27	ICSA	-.01203	.00189	.00200	.00145	.03084	-.12490
28	ICSAB	-.01139	.00178	.00676	.00171	.02229	-.14062
29	CCV	4.8343	4.8601	.19971	5.2782	4.8137	9.7132
30	CCB	.01260	.00008	.00775	.00049	.01123	.04738

#	Sample Name	V	Y	Zn	Zr
1	ICV	5.0092	.00804	4.8063	5.0714
2	ICB	.00235	.00037	-.00031	.00101
3	LLS	.10300	.00064	.01984	.02040
4	ICSA	.00227	.00750	-.00351	-.00238
5	ICSAB	.46616	.00774	.93383	-.00260
6	PREPBLKTJA	.00154	.00025	-.00013	-.00024
7	S98T003393_L	.19099	.00061	.03401	-.13352
8	S98T003393	.12359	.02304	.01522	.03836
9	S98T003393_D	.09438	.01372	.00821	-.00331
10	S98T003393_A	40.685	.07687	39.620	40.641
11	S98T003393_X	.97326	.12539	-.01510	.16560

Analysis Report

Averages

Mon 01-18-99 12:21:49 PM

page 6

#	Sample Name	V	Y	Zn	Zr
12	S98T003393_AX	4061.3	6.4117	3853.3	4064.3
13	CCV	5.0011	.00797	4.7899	5.0729
14	CCB	.00144	.00030	.00010	.00079
15	S98T003398_L	.32738	.06076	.26766	.11212
16	S98T003398	.14857	.05060	.26690	.06521
17	S98T003398_D	.27342	.07367	.27291	.11688
18	S98T003398_A	40.598	.13072	40.100	40.324
19	S98T003398_X	2.0080	.40587	.60569	.90432
20	S98T003398_AX	4051.5	6.6340	3912.1	4023.9
21	S98T003388_L	.41691	.04752	.02699	.16531
22	S98T003388	.16126	.03255	.00688	.04910
23	S98T003388_D	.31117	.06369	-.01063	.17813
24	S98T003388_A	40.344	.11259	40.167	39.972
25	S98T003388_X	2.6200	.53283	-.15596	1.3699
26	S98T003388_AX	4036.5	6.6551	3932.2	3996.9
27	ICSA	.00393	.00749	-.00353	-.00192
28	ICSAB	.46817	.00781	.93654	-.00179
29	CCV	5.0245	.00825	4.8496	5.0695
30	CCB	.00399	.00060	-.00002	.00131

JK *[Signature]*
01-18-99

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 7 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 2015 THRU 2379)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INDUCTIVELY COUPLED PLASMA/MASS SPECTROMETER ANALYSIS, ICP/MS WORKLIST 26708 THRU 27496		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26708

Analyst: bjg

Instrument: ICPMS1

Book#: 5859

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	MSACT TH232	QC	1.0e-2	9.61e-03	96.100 % Recovery	
1 ICV	0	MSACT NP237	QC	9.84e-3	1.32e-02	134.146 % Recovery	
1 ICV	0	MSACT PU239	QC	1.05e-2	1.46e-02	139.048 % Recovery	
1 ICV	0	MSACT 241	QC	n/a	2.19e-05	% Recovery	
2 ICB	0	MSACT1 PA231-A1	LIQUID	1	<1.54e-5	ug/mL	
2 ICB	0	MSACT1 PU239-A1	LIQUID	1	<1.75e-5	ug/mL	
2 ICB	0	MSACT1 PU240-A1	LIQUID	1	<1.02e-5	ug/mL	
2 ICB	0	MSACT1 PU242-A1	LIQUID	1	<2.39e-5	ug/mL	
2 ICB	0	MSACT1 241-A1	LIQUID	1	<1.58e-5	ug/mL	
2 ICB	0	MSACT1 243-A1	LIQUID	1	<1.44e-5	ug/mL	
2 ICB	0	MSACT1 244-A1	LIQUID	1	<1.22e-5	ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 PA231-A1	LIQUID	N/A <	6.253e-01	0.625	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 TH229-A1	LIQUID	N/A <	1.268e+00	1.268	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 TH230-A1	LIQUID	N/A <	8.695e-01	0.870	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 TH232-A1	LIQUID	N/A <	7.857e-01	0.786	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 241-A1	LIQUID	N/A <	6.408e-01	0.641	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 243-A1	LIQUID	N/A <	5.830e-01	0.583	ug/mL
3 SAMPLE	S98T002822 0 A	MSACT1 244-A1	LIQUID	N/A <	4.959e-01	0.496	ug/mL
4 DUP	S98T002822 0 A	MSACT1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
4 DUP	S98T002822 0 A	MSACT1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
4 DUP	S98T002822 0 A	MSACT1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
5 SAMPLE	S98T002823 0 A	MSACT1 PA231-A1	LIQUID	N/A <	6.253e-01	0.625	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 TH229-A1	LIQUID	N/A <	1.268e+00	1.268	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 TH230-A1	LIQUID	N/A <	8.695e-01	0.870	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 TH232-A1	LIQUID	N/A <	7.857e-01	0.786	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
5 SAMPLE	S98T002823 0 A	MSACT1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

2015

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 SAMPLE	S98T002823	0 A	OMSACT1	241-A1	LIQUID	N/A <	6.408e-01	0.641 ug/mL
5 SAMPLE	S98T002823	0 A	OMSACT1	243-A1	LIQUID	N/A <	5.830e-01	0.583 ug/mL
5 SAMPLE	S98T002823	0 A	OMSACT1	244-A1	LIQUID	N/A <	4.959e-01	0.496 ug/mL
6 DUP	S98T002823	0 A	OMSACT1	PA231-A1	LIQUID	<6.25e-1	<6.25e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	TH229-A1	LIQUID	<1.27e0	<1.27e0	RPD
6 DUP	S98T002823	0 A	OMSACT1	TH230-A1	LIQUID	<8.70e-1	<8.69e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	TH232-A1	LIQUID	<7.86e-1	<7.86e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	NP237-A1	LIQUID	<6.53e-1	<6.53e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	FU239-A1	LIQUID	<7.09e-1	<7.09e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	FU240-A1	LIQUID	<4.14e-1	<4.14e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	FU242-A1	LIQUID	<9.68e-1	<9.68e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	241-A1	LIQUID	<6.41e-1	<6.41e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	243-A1	LIQUID	<5.83e-1	<5.83e-1	RPD
6 DUP	S98T002823	0 A	OMSACT1	244-A1	LIQUID	<4.96e-1	<4.96e-1	RPD
7 SAMPLE	S98T002824	0 A	OMSACT1	PA231-A1	LIQUID	N/A <	6.253e-01	0.625 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	TH229-A1	LIQUID	N/A <	1.268e+00	1.268 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	TH230-A1	LIQUID	N/A <	8.695e-01	0.870 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	TH232-A1	LIQUID	N/A <	7.857e-01	0.786 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	NP237-A1	LIQUID	N/A <	6.534e-01	0.653 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	FU239-A1	LIQUID	N/A <	7.088e-01	0.709 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	FU240-A1	LIQUID	N/A <	4.140e-01	0.414 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	FU242-A1	LIQUID	N/A <	9.675e-01	0.968 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	241-A1	LIQUID	N/A <	6.408e-01	0.641 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	243-A1	LIQUID	N/A <	5.830e-01	0.583 ug/mL
7 SAMPLE	S98T002824	0 A	OMSACT1	244-A1	LIQUID	N/A <	4.959e-01	0.496 ug/mL
8 DUP	S98T002824	0 A	OMSACT1	PA231-A1	LIQUID	<6.25e-1	<6.25e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	TH229-A1	LIQUID	<1.27e0	<1.27e0	RPD
8 DUP	S98T002824	0 A	OMSACT1	TH230-A1	LIQUID	<8.70e-1	<8.69e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	TH232-A1	LIQUID	<7.86e-1	<7.86e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	NP237-A1	LIQUID	<6.53e-1	<6.53e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	FU239-A1	LIQUID	<7.09e-1	<7.09e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	FU240-A1	LIQUID	<4.14e-1	<4.14e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	FU242-A1	LIQUID	<9.68e-1	<9.68e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	241-A1	LIQUID	<6.41e-1	<6.41e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	243-A1	LIQUID	<5.83e-1	<5.83e-1	RPD
8 DUP	S98T002824	0 A	OMSACT1	244-A1	LIQUID	<4.96e-1	<4.96e-1	RPD
9 SAMPLE	S98T002825	0 A	OMSACT1	PA231-A1	LIQUID	N/A <	6.253e-01	0.625 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	TH229-A1	LIQUID	N/A <	1.268e+00	1.268 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	TH230-A1	LIQUID	N/A <	8.695e-01	0.870 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	TH232-A1	LIQUID	N/A <	7.857e-01	0.786 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	NP237-A1	LIQUID	N/A <	6.534e-01	0.653 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	FU239-A1	LIQUID	N/A <	7.088e-01	0.709 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	FU240-A1	LIQUID	N/A <	4.140e-01	0.414 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	FU242-A1	LIQUID	N/A <	9.675e-01	0.968 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	241-A1	LIQUID	N/A <	6.408e-01	0.641 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	243-A1	LIQUID	N/A <	5.830e-01	0.583 ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1	244-A1	LIQUID	N/A <	4.959e-01	0.496 ug/mL
10 DUP	S98T002825	0 A	OMSACT1	PA231-A1	LIQUID	<6.25e-1	<6.25e-1	RPD
10 DUP	S98T002825	0 A	OMSACT1	TH229-A1	LIQUID	<1.27e0	<1.27e0	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

2016

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
10 DUP	S98T002825	0 A	EMSACTAL	TH230-A1 LIQUID	<8.70e-1	<8.69e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	TH232-A1 LIQUID	<7.86e-1	<7.86e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	NP237-A1 LIQUID	<6.53e-1	<6.53e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	FU239-A1 LIQUID	<7.09e-1	<7.09e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	FU240-A1 LIQUID	<4.14e-1	<4.14e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	FU242-A1 LIQUID	<9.68e-1	<9.68e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	241-A1 LIQUID	<6.41e-1	<6.41e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	243-A1 LIQUID	<5.83e-1	<5.83e-1		RPD
10 DUP	S98T002825	0 A	EMSACTAL	244-A1 LIQUID	<4.96e-1	<4.96e-1		RPD
11 CCV	0	0	EMSACT	TH232 QC	1.0e-2	9.16e-03	91.600 % Recovery	
11 CCV	0	0	EMSACT	NP237 QC	9.84e-3	1.27e-02	129.065 % Recovery	
11 CCV	0	0	EMSACT	FU239 QC	1.05e-2	1.40e-02	133.333 % Recovery	
11 CCV	0	0	EMSACT	241 QC	n/a	1.87e-05	% Recovery	
12 CCB	0	0	EMSACTAL	PA231-A1 LIQUID	1	<1.54e-5		ug/mL
12 CCB	0	0	EMSACTAL	TH229-A1 LIQUID	1	<3.13e-5		ug/mL
12 CCB	0	0	EMSACTAL	TH230-A1 LIQUID	1	<2.15e-5		ug/mL
12 CCB	0	0	EMSACTAL	TH232-A1 LIQUID	1	<1.94e-5		ug/mL
12 CCB	0	0	EMSACTAL	FU239-A1 LIQUID	1	<1.75e-5		ug/mL
12 CCB	0	0	EMSACTAL	FU240-A1 LIQUID	1	<1.02e-5		ug/mL
12 CCB	0	0	EMSACTAL	FU242-A1 LIQUID	1	<2.39e-5		ug/mL
12 CCB	0	0	EMSACTAL	241-A1 LIQUID	1	<1.58e-5		ug/mL
12 CCB	0	0	EMSACTAL	243-A1 LIQUID	1	<1.44e-5		ug/mL
12 CCB	0	0	EMSACTAL	244-A1 LIQUID	1	<1.22e-5		ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	PA231-A1 LIQUID	N/A	< 6.253e-01	0.625	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	TH229-A1 LIQUID	N/A	< 1.268e+00	1.268	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	TH230-A1 LIQUID	N/A	< 8.695e-01	0.870	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	TH232-A1 LIQUID	N/A	< 7.857e-01	0.786	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	NP237-A1 LIQUID	N/A	< 6.534e-01	0.653	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	FU239-A1 LIQUID	N/A	< 7.088e-01	0.709	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	FU240-A1 LIQUID	N/A	< 4.140e-01	0.414	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	FU242-A1 LIQUID	N/A	< 9.675e-01	0.968	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	241-A1 LIQUID	N/A	< 6.408e-01	0.641	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	243-A1 LIQUID	N/A	< 5.830e-01	0.583	ug/mL
13 SAMPLE	S98T002827	0 A	EMSACTAL	244-A1 LIQUID	N/A	< 4.959e-01	0.496	ug/mL
14 DUP	S98T002827	0 A	EMSACTAL	PA231-A1 LIQUID	<6.25e-1	<6.25e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	TH229-A1 LIQUID	<1.27e0	<1.27e0		RPD
14 DUP	S98T002827	0 A	EMSACTAL	TH230-A1 LIQUID	<8.70e-1	<8.69e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	TH232-A1 LIQUID	<7.86e-1	<7.86e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	NP237-A1 LIQUID	<6.53e-1	<6.53e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	FU239-A1 LIQUID	<7.09e-1	<7.09e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	FU240-A1 LIQUID	<4.14e-1	<4.14e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	FU242-A1 LIQUID	<9.68e-1	<9.68e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	241-A1 LIQUID	<6.41e-1	<6.41e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	243-A1 LIQUID	<5.83e-1	<5.83e-1		RPD
14 DUP	S98T002827	0 A	EMSACTAL	244-A1 LIQUID	<4.96e-1	<4.96e-1		RPD
15 SPK-POST	S98T002827	0 A	EMSACTAL	TH232-A1 LIQUID	2.10e-2	1.71e-02	81.429 % Recovery	
15 SPK-POST	S98T002827	0 A	EMSACTAL	NP237-A1 LIQUID	1.68e-2	2.47e-02	147.024 % Recovery	
15 SPK-POST	S98T002827	0 A	EMSACTAL	FU239-A1 LIQUID	2.10e-2	2.71e-02	129.048 % Recovery	
15 SPK-POST	S98T002827	0 A	EMSACTAL	241-A1 LIQUID	n/a	4.30e-05	% Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2017

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
16 SAMPLE	S98T002940	0 A	OMSACTA1 PA231-A1	LIQUID	N/A	< 6.253e-01	0.625	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 TH229-A1	LIQUID	N/A	< 1.268e+00	1.268	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 TH230-A1	LIQUID	N/A	< 8.695e-01	0.870	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 TH232-A1	LIQUID	N/A	< 7.857e-01	0.786	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 NP237-A1	LIQUID	N/A	< 6.534e-01	0.653	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 PU239-A1	LIQUID	N/A	< 7.088e-01	0.709	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 PU240-A1	LIQUID	N/A	< 4.140e-01	0.414	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 PU242-A1	LIQUID	N/A	< 9.675e-01	0.968	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 241-A1	LIQUID	N/A	< 6.408e-01	0.641	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 243-A1	LIQUID	N/A	< 5.830e-01	0.583	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACTA1 244-A1	LIQUID	N/A	< 4.959e-01	0.496	ug/mL
17 DUP	S98T002940	0 A	OMSACTA1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
17 DUP	S98T002940	0 A	OMSACTA1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
17 DUP	S98T002940	0 A	OMSACTA1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
18 SPK-POST	S98T002940	0 A	OMSACTA1 TH232-A1	LIQUID	2.10e-2	1.82e-02	86.667 %	Recovery
18 SPK-POST	S98T002940	0 A	OMSACTA1 NP237-A1	LIQUID	1.68e-2	2.59e-02	154.167 %	Recovery
18 SPK-POST	S98T002940	0 A	OMSACTA1 PU239-A1	LIQUID	2.10e-2	2.86e-02	136.190 %	Recovery
18 SPK-POST	S98T002940	0 A	OMSACTA1 241-A1	LIQUID	n/a	4.53e-05		% Recovery
19 CCV		0	OMSACT TH232	QC	1.0e-2	9.46e-03	94.600 %	Recovery
19 CCV		0	OMSACT NP237	QC	9.84e-3	1.30e-02	132.114 %	Recovery
19 CCV		0	OMSACT PU239	QC	1.05e-2	1.43e-02	136.190 %	Recovery
19 CCV		0	OMSACT 241	QC	n/a	1.76e-05		% Recovery
20 CCB		0	OMSACTA1 PA231-A1	LIQUID	1	<1.54e-5		ug/mL
20 CCB		0	OMSACTA1 TH229-A1	LIQUID	1	<3.13e-5		ug/mL
20 CCB		0	OMSACTA1 TH230-A1	LIQUID	1	<2.15e-5		ug/mL
20 CCB		0	OMSACTA1 TH232-A1	LIQUID	1	<1.94e-5		ug/mL
20 CCB		0	OMSACTA1 PU239-A1	LIQUID	1	<1.75e-5		ug/mL
20 CCB		0	OMSACTA1 PU240-A1	LIQUID	1	<1.02e-5		ug/mL
20 CCB		0	OMSACTA1 PU242-A1	LIQUID	1	<2.39e-5		ug/mL
20 CCB		0	OMSACTA1 241-A1	LIQUID	1	<1.58e-5		ug/mL
20 CCB		0	OMSACTA1 243-A1	LIQUID	1	<1.44e-5		ug/mL
20 CCB		0	OMSACTA1 244-A1	LIQUID	1	<1.22e-5		ug/mL

Final page for worklist# 26708

Analyst Signature

Date

Analyst Signature

Date

Sal M. Parg 12/7/98

LABCORE Completed Worklist Report for Worklist# 26708

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or	Yield	Unit
				Reviewer	Signature	Date				

SP 12/7/98

LABCORE Completed Worklist Report for Worklist# 26708

Analyst: bjs/KST 8/12/7/98 Instrument: ICPMS1

Book# 5B59

Method: LA-506101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	MSACT TH232	QC	1.0e-2	9.61e-03	96.100 % Recovery	
1 ICV	0	MSACT NP237	QC	9.84e-3	1.32e-02	134.146 % Recovery	
1 ICV	0	MSACT PU239	QC	1.05e-2	1.46e-02	139.048 % Recovery	
1 ICV	0	MSACT 241	QC	8.91e-3	2.19e-05	0.246 % Recovery	
2 ICB	0	MSACT1 PA231-A1	LIQUID	1	<1.54e-5	ug/mL	
2 ICB	0	MSACT1 TH229-A1	LIQUID	1	<3.13e-5	ug/mL	
2 ICB	0	MSACT1 TH230-A1	LIQUID	1	<2.15e-5	ug/mL	
2 ICB	0	MSACT1 TH232-A1	LIQUID	1	<1.94e-5	ug/mL	
2 ICB	0	MSACT1 NP237-A1	LIQUID	1	<1.61e-5	ug/mL	
2 ICB	0	MSACT1 PU239-A1	LIQUID	1	<1.75e-5	ug/mL	
2 ICB	0	MSACT1 PU240-A1	LIQUID	1	<1.02e-5	ug/mL	
2 ICB	0	MSACT1 PU242-A1	LIQUID	1	<2.39e-5	ug/mL	
2 ICB	0	MSACT1 241-A1	LIQUID	1	<1.58e-5	ug/mL	
2 ICB	0	MSACT1 243-A1	LIQUID	1	<1.44e-5	ug/mL	
2 ICB	0	MSACT1 244-A1	LIQUID	1	<1.22e-5	ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 PA231-A1	LIQUID	N/A	< 6.253e-01	0.625 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 TH229-A1	LIQUID	N/A	< 1.268e+00	1.268 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 TH230-A1	LIQUID	N/A	< 8.695e-01	0.870 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 TH232-A1	LIQUID	N/A	< 7.857e-01	0.786 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 NP237-A1	LIQUID	N/A	< 6.534e-01	0.653 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 PU239-A1	LIQUID	N/A	< 7.088e-01	0.709 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 PU240-A1	LIQUID	N/A	< 4.140e-01	0.414 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 PU242-A1	LIQUID	N/A	< 9.675e-02	0.968 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 241-A1	LIQUID	N/A	< 6.408e-01	0.641 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 243-A1	LIQUID	N/A	< 5.830e-01	0.583 ug/mL	
3 SAMPLE	S98T002822 0 A	MSACT1 244-A1	LIQUID	N/A	< 4.959e-01	0.496 ug/mL	
4 DUP	S98T002822 0 A	MSACT1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 TH229-A1	LIQUID	<1.27e0	<1.27e0	RPD	
4 DUP	S98T002822 0 A	MSACT1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 241-A1	LIQUID	<6.41e-1	<6.41e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 243-A1	LIQUID	<5.83e-1	<5.83e-1	RPD	
4 DUP	S98T002822 0 A	MSACT1 244-A1	LIQUID	<4.96e-1	<4.96e-1	RPD	
5 SAMPLE	S98T002823 0 A	MSACT1 PA231-A1	LIQUID	N/A	< 6.253e-01	0.625 ug/mL	
5 SAMPLE	S98T002823 0 A	MSACT1 TH229-A1	LIQUID	N/A	< 1.268e+00	1.268 ug/mL	
5 SAMPLE	S98T002823 0 A	MSACT1 TH230-A1	LIQUID	N/A	< 8.695e-01	0.870 ug/mL	
5 SAMPLE	S98T002823 0 A	MSACT1 TH232-A1	LIQUID	N/A	< 7.857e-01	0.786 ug/mL	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 SAMPLE	S98T002823	0 A	OMSACTA1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 241-A1	LIQUID	N/A <	6.408e-01	0.641	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 243-A1	LIQUID	N/A <	5.830e-01	0.583	ug/mL
5 SAMPLE	S98T002823	0 A	OMSACTA1 244-A1	LIQUID	N/A <	4.959e-01	0.496	ug/mL
6 DUP	S98T002823	0 A	OMSACTA1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
6 DUP	S98T002823	0 A	OMSACTA1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
6 DUP	S98T002823	0 A	OMSACTA1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
7 SAMPLE	S98T002824	0 A	OMSACTA1 PA231-A1	LIQUID	N/A <	6.253e-01	0.625	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 TH229-A1	LIQUID	N/A <	1.268e+00	1.268	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 TH230-A1	LIQUID	N/A <	8.695e-01	0.870	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 TH232-A1	LIQUID	N/A <	7.857e-01	0.786	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 241-A1	LIQUID	N/A <	6.408e-01	0.641	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 243-A1	LIQUID	N/A <	5.830e-01	0.583	ug/mL
7 SAMPLE	S98T002824	0 A	OMSACTA1 244-A1	LIQUID	N/A <	4.959e-01	0.496	ug/mL
8 DUP	S98T002824	0 A	OMSACTA1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
8 DUP	S98T002824	0 A	OMSACTA1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
8 DUP	S98T002824	0 A	OMSACTA1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
9 SAMPLE	S98T002825	0 A	OMSACTA1 PA231-A1	LIQUID	N/A <	6.253e-01	0.625	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 TH229-A1	LIQUID	N/A <	1.268e+00	1.268	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 TH230-A1	LIQUID	N/A <	8.695e-01	0.870	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 TH232-A1	LIQUID	N/A <	7.857e-01	0.786	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACTA1 241-A1	LIQUID	N/A <	6.408e-01	0.641	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 SAMPLE	S98T002825	0 A	OMSACT1 243-A1	LIQUID	N/A <	5.830e-01	0.583	ug/mL
9 SAMPLE	S98T002825	0 A	OMSACT1 244-A1	LIQUID	N/A <	4.959e-01	0.496	ug/mL
10 DUP	S98T002825	0 A	OMSACT1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
10 DUP	S98T002825	0 A	OMSACT1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
10 DUP	S98T002825	0 A	OMSACT1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
11 CCV	0		OMSACT TH232	QC	1.0e-2	9.16e-03	91.600 % Recovery	
11 CCV	0		OMSACT NP237	QC	9.84e-3	1.27e-02	129.065 % Recovery	
11 CCV	0		OMSACT PU239	QC	1.05e-2	1.40e-02	133.333 % Recovery	
11 CCV	0		OMSACT 241	QC	8.91e-3	1.87e-05	0.210 % Recovery	
12 CCB	0		OMSACT1 PA231-A1	LIQUID	1	<1.54e-5		ug/mL
12 CCB	0		OMSACT1 TH229-A1	LIQUID	1	<3.13e-5		ug/mL
12 CCB	0		OMSACT1 TH230-A1	LIQUID	1	<2.15e-5		ug/mL
12 CCB	0		OMSACT1 TH232-A1	LIQUID	1	<1.94e-5		ug/mL
12 CCB	0		OMSACT1 NP237-A1	LIQUID	1	<1.61e-5		ug/mL
12 CCB	0		OMSACT1 PU239-A1	LIQUID	1	<1.75e-5		ug/mL
12 CCB	0		OMSACT1 PU240-A1	LIQUID	1	<1.02e-5		ug/mL
12 CCB	0		OMSACT1 PU242-A1	LIQUID	1	<2.39e-5		ug/mL
12 CCB	0		OMSACT1 241-A1	LIQUID	1	<1.58e-5		ug/mL
12 CCB	0		OMSACT1 243-A1	LIQUID	1	<1.44e-5		ug/mL
12 CCB	0		OMSACT1 244-A1	LIQUID	1	<1.22e-5		ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 PA231-A1	LIQUID	N/A <	6.253e-01	0.625	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 TH229-A1	LIQUID	N/A <	1.268e+00	1.268	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 TH230-A1	LIQUID	N/A <	8.695e-01	0.870	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 TH232-A1	LIQUID	N/A <	7.857e-01	0.786	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 NP237-A1	LIQUID	N/A <	6.534e-01	0.653	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 PU239-A1	LIQUID	N/A <	7.088e-01	0.709	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 PU240-A1	LIQUID	N/A <	4.140e-01	0.414	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 PU242-A1	LIQUID	N/A <	9.675e-01	0.968	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 241-A1	LIQUID	N/A <	6.408e-01	0.641	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 243-A1	LIQUID	N/A <	5.830e-01	0.583	ug/mL
13 SAMPLE	S98T002827	0 A	OMSACT1 244-A1	LIQUID	N/A <	4.959e-01	0.496	ug/mL
14 DUP	S98T002827	0 A	OMSACT1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
14 DUP	S98T002827	0 A	OMSACT1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
14 DUP	S98T002827	0 A	OMSACT1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

2022

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
14 DUP	S98T002827	0 A	OMSACT1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
15 SPK-POST	S98T002827	0 A	OMSACT1 PA231-A1	LIQUID	n/a	3.43e-05		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 TH229-A1	LIQUID	n/a	-7.11e-6		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 TH230-A1	LIQUID	n/a	-1.51e-6		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 TH232-A1	LIQUID	2.10e-2	1.71e-02	81.429	% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 NP237-A1	LIQUID	1.68e-2	2.47e-02	147.024	% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 PU239-A1	LIQUID	2.10e-2	2.71e-02	129.048	% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 PU240-A1	LIQUID	n/a	6.56e-04		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 PU242-A1	LIQUID	n/a	-2.63e-6		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 241-A1	LIQUID	1.78e-2	4.30e-05	0.242	% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 243-A1	LIQUID	n/a	-1.39e-6		% Recovery
15 SPK-POST	S98T002827	0 A	OMSACT1 244-A1	LIQUID	?	-4.13e-6		% Recovery
16 SAMPLE	S98T002940	0 A	OMSACT1 PA231-A1	LIQUID	N/A	< 6.253e-01	0.625	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 TH229-A1	LIQUID	N/A	< 1.268e+00	1.268	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 TH230-A1	LIQUID	N/A	< 8.695e-01	0.870	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 TH232-A1	LIQUID	N/A	< 7.857e-01	0.786	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 NP237-A1	LIQUID	N/A	< 6.534e-01	0.653	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 PU239-A1	LIQUID	N/A	< 7.088e-01	0.709	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 PU240-A1	LIQUID	N/A	< 4.140e-01	0.414	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 PU242-A1	LIQUID	N/A	< 9.675e-01	0.968	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 241-A1	LIQUID	N/A	< 6.408e-01	0.641	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 243-A1	LIQUID	N/A	< 5.830e-01	0.583	ug/mL
16 SAMPLE	S98T002940	0 A	OMSACT1 244-A1	LIQUID	N/A	< 4.959e-01	0.496	ug/mL
17 DUP	S98T002940	0 A	OMSACT1 PA231-A1	LIQUID	<6.25e-1	<6.25e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 TH229-A1	LIQUID	<1.27e0	<1.27e0		RPD
17 DUP	S98T002940	0 A	OMSACT1 TH230-A1	LIQUID	<8.70e-1	<8.69e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 TH232-A1	LIQUID	<7.86e-1	<7.86e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 NP237-A1	LIQUID	<6.53e-1	<6.53e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 PU239-A1	LIQUID	<7.09e-1	<7.09e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 PU240-A1	LIQUID	<4.14e-1	<4.14e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 PU242-A1	LIQUID	<9.68e-1	<9.68e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 241-A1	LIQUID	<6.41e-1	<6.41e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 243-A1	LIQUID	<5.83e-1	<5.83e-1		RPD
17 DUP	S98T002940	0 A	OMSACT1 244-A1	LIQUID	<4.96e-1	<4.96e-1		RPD
18 SPK-POST	S98T002940	0 A	OMSACT1 PA231-A1	LIQUID	n/a	3.16e-05		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 TH229-A1	LIQUID	n/a	-7.51e-6		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 TH230-A1	LIQUID	n/a	-3.78e-6		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 TH232-A1	LIQUID	2.10e-2	1.82e-02	86.667	% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 NP237-A1	LIQUID	1.68e-2	2.59e-02	154.167	% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 PU239-A1	LIQUID	2.10e-2	2.86e-02	136.190	% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 PU240-A1	LIQUID	n/a	7.02e-04		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 PU242-A1	LIQUID	n/a	-3.21e-6		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 241-A1	LIQUID	1.78e-2	4.53e-05	0.254	% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 243-A1	LIQUID	n/a	-3.47e-6		% Recovery
18 SPK-POST	S98T002940	0 A	OMSACT1 244-A1	LIQUID	?	-4.51e-6		% Recovery
19 CCV		0	OMSACT TH232	QC	1.0e-2	9.46e-03	94.600	% Recovery
19 CCV		0	OMSACT NP237	QC	9.84e-3	1.30e-02	132.114	% Recovery
19 CCV		0	OMSACT PU239	QC	1.05e-2	1.43e-02	136.190	% Recovery
19 CCV		0	OMSACT 241	QC	8.91e-3	1.76e-05	0.198	% Recovery

Units shown for QC (B/L/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26708

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
20 CCB	0	EMSACTA1 PA231-A1	LIQUID	1	<1.54e-5		ug/mL
20 CCB	0	EMSACTA1 TH229-A1	LIQUID	1	<3.13e-5		ug/mL
20 CCB	0	EMSACTA1 TH230-A1	LIQUID	1	<2.15e-5		ug/mL
20 CCB	0	EMSACTA1 TH232-A1	LIQUID	1	<1.94e-5		ug/mL
20 CCB	0	EMSACTA1 NP237-A1	LIQUID	1	<1.61e-5		ug/mL
20 CCB	0	EMSACTA1 PU239-A1	LIQUID	1	<1.75e-5		ug/mL
20 CCB	0	EMSACTA1 PU240-A1	LIQUID	1	<1.02e-5		ug/mL
20 CCB	0	EMSACTA1 PU242-A1	LIQUID	1	<2.39e-5		ug/mL
20 CCB	0	EMSACTA1 241-A1	LIQUID	1	<1.58e-5		ug/mL
20 CCB	0	EMSACTA1 243-A1	LIQUID	1	<1.44e-5		ug/mL
20 CCB	0	EMSACTA1 244-A1	LIQUID	1	<1.22e-5		ug/mL

Final page for worklist# 26708

for: *Sal M. Bag*

KST

12/01/98

Analyst Signature

Date

Analyst Signature

Date

Sal M. Bag

Reviewer Signature

12/7/98

Date

10/22/98 12:07
A-0004-1

Use ac1201a.txt

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26708

Analyst: KSP Instrument: ICPMS1 Book# SB59

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICB			@MSACTA1	LIQUID		
3 SAMPLE	S98T002822 0 A		@MSACTA1	LIQUID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
4 DUP	S98T002822 0 A		@MSACTA1	LIQUID		
5 SAMPLE	S98T002823 0 A		@MSACTA1	LIQUID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
6 DUP	S98T002823 0 A		@MSACTA1	LIQUID		
7 SAMPLE	S98T002824 0 A		@MSACTA1	LIQUID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
8 DUP	S98T002824 0 A		@MSACTA1	LIQUID		
9 SAMPLE	S98T002825 0 A		@MSACTA1	LIQUID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
10 DUP	S98T002825 0 A		@MSACTA1	LIQUID		
11 CCV			@MSACT	QC		
12 CCB			@MSACTA1	LIQUID		
13 SAMPLE	S98T002827 0 A		@MSACTA1	LIQUID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
14 DUP	S98T002827 0 A		@MSACTA1	LIQUID		

Data Entry Comments:

Updated & Validated by:
Sam H. Pang

12/7/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

10/22/98 12:07

Page 2

A-0004-1

LABCORE Data Entry Template for Worklist# 26708

S Type	Sample#	R A	Test	Matrix	Group#	Project
15 SPK- <i>post</i> <i>8/12/98</i>	S98T002827	0 A	@MSACTAL	LIQUID		
16 SAMPLE	S98T002940	0 A	@MSACTAL	LIQUID	98000490	BY-105 (2)
Analytes Requested: 241-A1, 243-A1, 244-A1, NF237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
17 DUP	S98T002940	0 A	@MSACTAL	LIQUID		
18 SPK- <i>post</i> <i>8/12/98</i>	S98T002940	0 A	@MSACTAL	LIQUID		
19 CCV			@MSACT	QC		
20 CCB			@MSACTAL	LIQUID		

Final page for worklist # 26708

K. S. Remade 12-01-98
 Analyst Signature Date

_____
 Analyst Signature Date

BY-105
S98T002822 -23, 24, 25, 27, 29 .1-10-.025-10
Scans 98002822

23
24
25
27
2940

spikes .04ml 4B59

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE

Sample Name: Sample-0

Operator:

Comment:

Run Time: 12/01/98 13:20

Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/s
Avg	135800.
Stddev	915.
%RSD	.6739

#1	137000.
#2	135900.
#3	135700.
#4	134000.
#5	135900.
#6	135100.
#7	136800.
#8	136200.
#9	136600.
#10	135000.

Ks Phomdike 12-01-98
File ac1201.txt

BY-LQS

8987002822
2823
2824
2825
2827
2940

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2027 TO 2035.

Method: ACTINIDE Sample Name: cal blank

Operator: kjt

Comment: background

Run Time: 12/01/98 13:23 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4241	.4181	.4533	.5162	.4200
Stddev	.0535	.0408	.0443	.0235	.0323
%RSD	12.62	9.765	9.779	4.551	7.701

#1	.4925	.4053	.4925	.5143	.4184
#2	.4169	.4783	.4344	.5353	.3993
#3	.4089	.4397	.4704	.4792	.4397
#4	.3490	.3803	.3848	.5145	.3803
#5	.4534	.3867	.4845	.5375	.4623

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4263	.4264	.4715	.4734	.4858
Stddev	.0366	.0444	.0390	.1003	.0561
%RSD	8.585	10.41	8.271	21.19	11.55

#1	.4576	.3792	.4620	.3443	.5797
#2	.3905	.3949	.4476	.5748	.4520
#3	.3825	.4397	.4265	.4089	.4968
#4	.4474	.4250	.5234	.4693	.4474
#5	.4534	.4934	.4979	.5690	.4534

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4752	.4427
Stddev	.0634	.0632
%RSD	13.34	14.28

#1	.4271	.5448
#2	.4037	.4256
#3	.4660	.3913
#4	.5279	.3937
#5	.5512	.4579

Sample Name: cal blank Run Time: 12/01/98 13:23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	75.551
Stddev	.793
%RSD	1.0490

#1	76.481
#2	75.965
#3	75.815
#4	74.507
#5	74.988

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.010 ppm cal

Run Time: 12/01/98 13:28 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4193	.4477	160.3	240.8	.7514
Stddev	.0436	.0216	5.5	7.1	.0602
%RSD	10.40	4.816	3.444	2.937	8.013

#1	.4566	.4566	152.2	235.3	.6606
#2	.3713	.4630	157.2	236.2	.7338
#3	.4467	.4692	165.6	252.2	.7670
#4	.3722	.4209	164.1	243.1	.8242
#5	.4500	.4285	162.6	237.2	.7714

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.005	.5106	.4459	.4283	.4592
Stddev	.063	.0632	.0252	.0357	.0293
%RSD	6.257	12.37	5.643	8.335	6.386

#1	1.046	.5497	.4699	.3901	.4256
#2	1.009	.4368	.4368	.4455	.4455
#3	.8978	.5188	.4331	.4376	.4963
#4	1.019	.5893	.4741	.4741	.4830
#5	1.054	.4585	.4157	.3943	.4457

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4593	.4826
Stddev	.0929	.0615
%RSD	20.23	12.74

#1	.4832	.3768
#2	.5067	.5067
#3	.3790	.5279
#4	.5760	.4830
#5	.3514	.5185

HNF-1663 REV. 0

Sample Name: cal std Run Time: 12/01/98 13:28

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	75.680
Stddev	1.458
%RSD	1.9270

#1	75.189
#2	76.317
#3	73.882
#4	75.229
#5	77.784

Method: ACTINIDE Sample Name: cal std

Operator: Kjt

Comment: 0.100 ppm

Run Time: 12/01/98 13:33 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4275	.6386	1584.	2343.	3.944
Stddev	.0687	.0267	68.	44.	.195
%RSD	16.07	4.188	4.304	1.889	4.938
#1	.4202	.6175	1483.	2279.	3.632
#2	.4817	.6541	1573.	2356.	3.973
#3	.4023	.6035	1654.	2349.	4.172
#4	.3299	.6513	1570.	2331.	3.963
#5	.5031	.6666	1640.	2402.	3.978

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.332	1.138	.4441	.4706	.4315
Stddev	.190	.094	.0575	.0518	.0400
%RSD	2.996	8.241	12.95	11.00	9.279
#1	6.007	1.132	.4331	.4888	.4116
#2	6.382	1.264	.4110	.4375	.3977
#3	6.341	1.063	.3848	.5466	.4854
#4	6.487	1.195	.5356	.4113	.4627
#5	6.442	1.036	.4558	.4687	.3999

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3866	.4493
Stddev	.0389	.0485
%RSD	10.07	10.79
#1	.3173	.5274
#2	.3977	.4066
#3	.4067	.4111
#4	.4070	.4499
#5	.4042	.4515

Sample Name: cal std Run Time: 12/01/98 13:33

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	76.939
Stddev	1.061
%RSD	1.3796

#1	77.737
#2	75.426
#3	76.224
#4	77.799
#5	77.512

Method: ACTINIDE Sample Name: icv

Operator: kjt

Comment: msact lcs

Run Time: 12/01/98 13:42 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3468	.3766	154.1	17.53	.6627
Stddev	.0277	.0605	3.2	.49	.0702
%RSD	7.978	16.07	2.061	2.819	10.60

#1	.3270	.4359	152.4	18.20	.5772
#2	.3491	.3075	156.4	17.41	.6940
#3	.3115	.4153	154.4	17.37	.6188
#4	.3767	.4094	149.7	16.88	.6632
#5	.3696	.3150	157.7	17.77	.7603

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	300.3	370.1	9.922	.3729	.3817
Stddev	5.9	5.2	.521	.0221	.0434
%RSD	1.978	1.409	5.256	5.918	11.37

#1	301.4	362.1	9.292	.3794	.3633
#2	309.5	374.9	10.17	.3615	.3491
#3	299.2	372.2	10.22	.3904	.3904
#4	293.1	367.9	9.445	.3930	.3521
#5	298.5	373.6	10.48	.3402	.4536

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.088	.3594
Stddev	.048	.0321
%RSD	4.429	8.923

#1	1.017	.3310
#2	1.126	.3740
#3	1.071	.3821
#4	1.085	.3193
#5	1.138	.3906

Sample Name: icv Run Time: 12/01/98 13:42

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.767
Stddev	1.250
%RSD	1.5478

#1	82.579
#2	80.215
#3	80.267
#4	81.418
#5	79.356

Method: ACTINIDE Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 12/01/98 13:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3460	.3714	.5848	.5149	.3465
Stddev	.0423	.0433	.0516	.0556	.0394
%RSD	12.23	11.66	8.819	10.80	11.38

#1	.2953	.4261	.6665	.5864	.3754
#2	.3303	.3633	.5284	.4583	.3344
#3	.3795	.3183	.5957	.4774	.3060
#4	.3990	.4030	.5642	.4917	.3990
#5	.3258	.3464	.5691	.5609	.3176

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4179	.5001	.4054	.4032	.3632
Stddev	.0474	.0615	.0343	.0423	.0337
%RSD	11.35	12.30	8.464	10.48	9.276

#1	.4893	.5906	.3459	.3586	.3965
#2	.4170	.5202	.4129	.4335	.3716
#3	.3754	.4448	.4284	.4570	.3917
#4	.3748	.5038	.4111	.3663	.3345
#5	.4330	.4413	.4289	.4000	.3217

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3866	.4038
Stddev	.0441	.0560
%RSD	11.41	13.87

#1	.4345	.3206
#2	.4211	.4129
#3	.3795	.4570
#4	.3224	.3789
#5	.3753	.4495

HNF-1663 REV.0

Sample Name: icb Run Time: 12/01/98 13:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.996
Stddev	1.361
%RSD	1.6809

#1	79.019
#2	80.739
#3	81.692
#4	82.706
#5	80.825

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002822

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 13:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3496	.3760	.4916	.8936	.3719
Stddev	.0436	.0359	.0363	.0873	.0646
%RSD	12.47	9.556	7.374	9.767	17.38

#1	.4117	.3235	.5419	.9536	.4747
#2	.3076	.3540	.5057	.8766	.3034
#3	.3086	.3943	.4929	.9902	.3729
#4	.3668	.4039	.4451	.7625	.3338
#5	.3533	.4043	.4724	.8853	.3746

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3826	.5189	.4151	.3676	.3887
Stddev	.0469	.0511	.0348	.0189	.0503
%RSD	12.25	9.844	8.391	5.144	12.95

#1	.4495	.5041	.3907	.3361	.3613
#2	.3961	.5689	.4298	.3666	.4172
#3	.3215	.5744	.4415	.3858	.4544
#4	.3627	.4575	.3668	.3751	.3874
#5	.3831	.4895	.4469	.3746	.3235

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4005	.3845
Stddev	.0246	.0692
%RSD	6.152	18.00

#1	.4117	.4873
#2	.4214	.3034
#3	.4201	.4115
#4	.3833	.3668
#5	.3660	.3533

Sample Name: s98t002822 Run Time: 12/01/98 13:57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	79.081
Stddev	1.185
%RSD	1.4978

#1	79.351
#2	79.098
#3	77.765
#4	80.876
#5	78.316

Method: ACTINIDE Sample Name: s98t002822_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:03 Type: Unk

Mode: IR

Corr.Fact: 1.000030

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3833	.3711	.4594	.8630	.3746
Stddev	.0396	.0278	.0421	.0517	.0150
%RSD	10.34	7.482	9.163	5.986	4.007

#1	.3681	.3596	.4104	.8589	.3850
#2	.3203	.3801	.4442	.9097	.3545
#3	.4036	.4079	.4423	.8502	.3736
#4	.4144	.3324	.4792	.9109	.3670
#5	.4098	.3757	.5208	.7855	.3927

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3727	.4600	.4064	.4533	.3925
Stddev	.0372	.0361	.0570	.0224	.0128
%RSD	9.982	7.838	14.01	4.936	3.265

#1	.4358	.4569	.3300	.4739	.3977
#2	.3673	.4997	.4527	.4570	.4100
#3	.3693	.4895	.3607	.4466	.3864
#4	.3411	.4101	.4403	.4706	.3756
#5	.3500	.4440	.4482	.4183	.3927

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3915	.3890
Stddev	.0740	.0513
%RSD	18.91	13.18

#1	.4189	.4062
#2	.3075	.4485
#3	.3650	.4122
#4	.3626	.3152
#5	.5037	.3628

Sample Name: s98t002822_d Run Time: 12/01/98 14:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	77.951
Stddev	.585
%RSD	.75049

#1	78.784
#2	78.044
#3	77.631
#4	77.212
#5	78.086

Method: ACTINIDE Sample Name: s98t002823

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:07 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3507	.3544	.4319	.7795	.3627
Stddev	.0410	.0224	.0442	.0749	.0293
%RSD	11.68	6.307	10.22	9.606	8.083

#1	.3412	.3497	.4971	.8173	.3497
#2	.3541	.3288	.4047	.7419	.4131
#3	.2865	.3653	.3861	.6809	.3446
#4	.3893	.3417	.4542	.7786	.3634
#5	.3823	.3867	.4174	.8788	.3427

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4083	.4731	.3899	.3519	.4090
Stddev	.0346	.0303	.0279	.0327	.0308
%RSD	8.472	6.403	7.161	9.299	7.538

#1	.4297	.4803	.3497	.3581	.3792
#2	.4257	.4595	.4131	.3330	.3752
#3	.4235	.4276	.4193	.3280	.4318
#4	.4153	.4974	.3807	.4066	.4153
#5	.3471	.5009	.3867	.3339	.4438

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3755	.3925
Stddev	.0120	.0300
%RSD	3.194	7.647

#1	.3960	.3539
#2	.3667	.3794
#3	.3695	.4359
#4	.3763	.3936
#5	.3691	.3998

Sample Name: s98t002823 Run Time: 12/01/98 14:07

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	78.283
Stddev	1.783
%RSD	2.2775

#1	79.123
#2	79.077
#3	80.290
#4	77.060
#5	75.864

Method: ACTINIDE Sample Name: s98t002823_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:12 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3465	.3765	.4767	.7679	.4017
Stddev	.0258	.0196	.0781	.0880	.0326
%RSD	7.450	5.207	16.38	11.45	8.125
#1	.3273	.4060	.3853	.6421	.4350
#2	.3683	.3725	.5504	.7118	.3601
#3	.3294	.3843	.4054	.8531	.4350
#4	.3270	.3647	.5491	.8216	.3898
#5	.3803	.3553	.4932	.8108	.3887

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3547	.4518	.3734	.3726	.4049
Stddev	.0484	.0384	.0439	.0297	.0510
%RSD	13.65	8.503	11.75	7.982	12.59
#1	.3439	.4557	.3811	.3604	.3853
#2	.3683	.4594	.3642	.3269	.4594
#3	.3041	.5068	.4434	.3928	.4223
#4	.3270	.4359	.3270	.3815	.3270
#5	.4305	.4012	.3511	.4012	.4305

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4133	.3875
Stddev	.0231	.0221
%RSD	5.592	5.702
#1	.4226	.3977
#2	.4056	.3559
#3	.3801	.4054
#4	.4150	.3731
#5	.4430	.4054

Sample Name: s98t002823_d Run Time: 12/01/98 14:12

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	79.841
Stddev	.674
%RSD	.84462

#1	80.460
#2	80.544
#3	78.929
#4	79.520
#5	79.752

Method: ACTINIDE Sample Name: s98t002824

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3405	.4020	.4281	.8429	.3555
Stddev	.0430	.0366	.0292	.0599	.0225
%RSD	12.64	9.100	6.814	7.110	6.315

#1	.3090	.3578	.3944	.8742	.3904
#2	.2877	.4480	.4357	.7728	.3288
#3	.3671	.3925	.4726	.8609	.3460
#4	.3945	.3817	.4242	.9162	.3563
#5	.3439	.4299	.4135	.7902	.3562

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3790	.4300	.3946	.3725	.4190
Stddev	.0652	.0786	.0136	.0474	.0332
%RSD	17.20	18.27	3.436	12.71	7.935

#1	.3212	.4676	.4066	.3741	.3741
#2	.3206	.3576	.3782	.3946	.4480
#3	.3798	.5444	.4051	.3629	.4515
#4	.3945	.4199	.3817	.3011	.4242
#5	.4791	.3603	.4013	.4299	.3972

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3608	.3895
Stddev	.0523	.0437
%RSD	14.49	11.23

#1	.2928	.4269
#2	.4193	.3535
#3	.4051	.4389
#4	.3308	.3393
#5	.3562	.3890

Sample Name: s98t002824 Run Time: 12/01/98 14:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.412
Stddev	1.522
%RSD	1.8924

#1	81.978
#2	81.095
#3	78.989
#4	78.588
#5	81.409

Method: ACTINIDE Sample Name: s98t002824_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:22 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3142	.3667	.4123	.8161	.3806
Stddev	.0201	.0366	.0575	.1221	.0292
%RSD	6.388	9.986	13.95	14.96	7.669

#1	.3334	.3575	.4097	.9439	.3655
#2	.3195	.3687	.4915	.7742	.3809
#3	.2947	.3431	.4360	.7589	.3431
#4	.3321	.4282	.3882	.6643	.3922
#5	.2915	.3361	.3361	.9394	.4211

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3685	.4130	.4024	.3645	.3782
Stddev	.0291	.0245	.0247	.0265	.0172
%RSD	7.885	5.942	6.146	7.265	4.549

#1	.3615	.4298	.3735	.3293	.3494
#2	.3523	.4260	.4096	.3891	.3932
#3	.4117	.4077	.3795	.3431	.3835
#4	.3361	.3722	.4202	.3802	.3882
#5	.3806	.4292	.4292	.3806	.3766

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4017	.3798
Stddev	.0397	.0309
%RSD	9.877	8.135

#1	.3414	.3936
#2	.4219	.4137
#3	.4481	.3391
#4	.3922	.3962
#5	.4049	.3563

Sample Name: s98t002824_d Run Time: 12/01/98 14:22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	82.513
Stddev	.738
%RSD	.89433

#1	82.993
#2	81.376
#3	82.575
#4	83.298
#5	82.324

Method: ACTINIDE Sample Name: s98t002825

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3511	.3801	.4516	1.184	.3574
Stddev	.0468	.0415	.0120	.070	.0487
%RSD	13.34	10.93	2.650	5.935	13.62

#1	.3440	.3906	.4443	1.079	.3261
#2	.3292	.3373	.4389	1.158	.4145
#3	.4188	.4228	.4662	1.233	.3912
#4	.3699	.4141	.4623	1.258	.3618
#5	.2936	.3355	.4461	1.193	.2936

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3498	.3696	.3835	.3923	.3800
Stddev	.0430	.0516	.0505	.0363	.0356
%RSD	12.30	13.97	13.16	9.249	9.360

#1	.3046	.2831	.3153	.3368	.4049
#2	.3820	.3780	.3820	.4308	.3292
#3	.3754	.4149	.4504	.3793	.3556
#4	.3859	.4020	.3618	.4141	.4061
#5	.3012	.3698	.4080	.4003	.4041

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3686	.3689
Stddev	.0152	.0148
%RSD	4.125	4.014

#1	.3727	.3512
#2	.3820	.3902
#3	.3595	.3595
#4	.3819	.3699
#5	.3470	.3736

Sample Name: s98t002825 Run Time: 12/01/98 14:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	85.949
Stddev	4.455
%RSD	5.1830

#1	93.021
#2	82.019
#3	84.365
#4	82.912
#5	87.428

Method: ACTINIDE Sample Name: s98t002825_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4285	.3929	.4025	1.222	.3904
Stddev	.0561	.0216	.0488	.104	.0370
%RSD	13.09	5.486	12.11	8.484	9.466

#1	.3492	.3923	.3845	1.303	.3610
#2	.4231	.3678	.4627	1.127	.3915
#3	.4164	.4243	.4441	1.348	.3529
#4	.5038	.4007	.3729	1.115	.4007
#5	.4498	.3794	.3481	1.217	.4459

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3502	.4093	.3889	.4062	.4039
Stddev	.0326	.0297	.0379	.0616	.0352
%RSD	9.313	7.264	9.744	15.16	8.727

#1	.3923	.4551	.3884	.4512	.4080
#2	.3519	.4113	.3796	.4904	.3875
#3	.3688	.4084	.4362	.3807	.4600
#4	.3094	.3729	.3332	.3491	.3650
#5	.3286	.3990	.4068	.3599	.3990

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3835	.4141
Stddev	.0786	.0404
%RSD	20.49	9.758

#1	.3767	.3884
#2	.3836	.4706
#3	.5115	.4164
#4	.3015	.3650
#5	.3442	.4303

Sample Name: s98t002825_d Run Time: 12/01/98 14:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.511
Stddev	.542
%RSD	.64153

#1	84.959
#2	84.292
#3	84.063
#4	84.029
#5	85.214

Method: ACTINIDE Sample Name: ccv

Operator: kjt

Comment: msact lcs

Run Time: 12/01/98 14:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3043	.3232	146.9	17.00	.6478
Stddev	.0379	.0408	5.4	.51	.0563
%RSD	12.44	12.62	3.710	2.994	8.692
#1	.2791	.2675	140.9	16.67	.7444
#2	.3187	.3148	149.3	17.40	.6256
#3	.3110	.3147	141.2	16.27	.6479
#4	.2570	.3400	152.7	17.42	.6168
#5	.3554	.3791	150.3	17.23	.6042

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	288.0	355.9	9.575	.3774	.3558
Stddev	10.0	11.4	.509	.0173	.0287
%RSD	3.461	3.191	5.315	4.584	8.055
#1	283.5	347.3	9.103	.3683	.3373
#2	291.8	361.5	9.719	.3935	.4053
#3	273.0	341.1	8.975	.3629	.3369
#4	298.6	368.8	10.07	.3638	.3440
#5	293.0	361.0	10.00	.3988	.3554

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.000	.3389
Stddev	.035	.0327
%RSD	3.465	9.642
#1	1.008	.3489
#2	1.019	.3463
#3	.9627	.2962
#4	1.044	.3203
#5	.9675	.3830

Sample Name: ccv Run Time: 12/01/98 14:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	85.887
Stddev	2.409
%RSD	2.8048

#1	85.978
#2	84.717
#3	90.028
#4	84.301
#5	84.413

Method: ACTINIDE Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 12/01/98 14:43 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3061	.3320	.4867	.4626	.3279
Stddev	.0244	.0273	.0539	.0476	.0116
%RSD	7.967	8.218	11.07	10.30	3.540

#1	.2919	.2994	.4715	.4191	.3218
#2	.2965	.3321	.5575	.5337	.3321
#3	.2800	.3589	.4614	.4811	.3116
#4	.3223	.3105	.5228	.4206	.3420
#5	.3399	.3590	.4201	.4583	.3322

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3672	.6558	.3676	.3819	.3561
Stddev	.0489	.0451	.0534	.0277	.0330
%RSD	13.30	6.870	14.53	7.244	9.272

#1	.3106	.6212	.3293	.4078	.3068
#2	.3400	.7156	.3242	.3361	.3677
#3	.4417	.6586	.4378	.3983	.3471
#4	.3734	.6800	.3341	.3813	.3616
#5	.3704	.6034	.4124	.3857	.3972

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3462	.3823
Stddev	.0406	.0468
%RSD	11.71	12.24

#1	.3106	.3405
#2	.3163	.3677
#3	.3983	.3392
#4	.3813	.4324
#5	.3246	.4315

Sample Name: ccb Run Time: 12/01/98 14:43

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	85.998
Stddev	2.099
%RSD	2.4403

#1	89.079
#2	84.310
#3	84.520
#4	84.799
#5	87.284

Method: ACTINIDE Sample Name: s98t002827

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:47 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3242	.3251	.4449	.5964	.3264
Stddev	.0252	.0320	.0392	.0686	.0453
%RSD	7.788	9.854	8.804	11.50	13.88

#1	.3191	.2994	.5082	.6934	.4018
#2	.3684	.3010	.4040	.6060	.3327
#3	.3158	.3718	.4397	.5157	.3118
#4	.3054	.3450	.4243	.6187	.2935
#5	.3121	.3081	.4481	.5482	.2921

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3483	.5307	.3638	.3582	.3709
Stddev	.0518	.0663	.0259	.0320	.0448
%RSD	14.88	12.50	7.116	8.939	12.07

#1	.2639	.5909	.3624	.4058	.4058
#2	.3327	.4991	.3763	.3288	.3763
#3	.3878	.4517	.3358	.3438	.2958
#4	.3768	.5037	.4006	.3371	.4045
#5	.3801	.6082	.3441	.3761	.3721

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3552	.3791
Stddev	.0288	.0386
%RSD	8.098	10.17

#1	.3270	.3664
#2	.3882	.3882
#3	.3838	.4317
#4	.3331	.3252
#5	.3441	.3841

Sample Name: s98t002827 Run Time: 12/01/98 14:47

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	83.901
Stddev	.550
%RSD	.65570

#1	84.612
#2	84.153
#3	83.384
#4	84.050
#5	83.308

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002827_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:52 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3456	.2943	.4297	.5775	.3552
Stddev	.0299	.0204	.0326	.0519	.0223
%RSD	8.657	6.943	7.583	8.978	6.275

#1	.3582	.2803	.4010	.5101	.3621
#2	.3489	.3018	.4155	.6272	.3371
#3	.3858	.3262	.4733	.6165	.3818
#4	.3276	.2877	.4036	.5354	.3276
#5	.3073	.2753	.4549	.5986	.3671

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3419	.5000	.3576	.3567	.3569
Stddev	.0344	.0601	.0441	.0131	.0497
%RSD	10.07	12.02	12.33	3.683	13.93

#1	.3387	.4594	.3115	.3660	.2803
#2	.2901	.5566	.3881	.3448	.4194
#3	.3341	.4176	.4176	.3500	.3580
#4	.3756	.5154	.3436	.3476	.3676
#5	.3711	.5507	.3272	.3751	.3591

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3626	.3409
Stddev	.0374	.0199
%RSD	10.32	5.828

#1	.2998	.3193
#2	.3606	.3685
#3	.3938	.3540
#4	.3716	.3316
#5	.3871	.3312

Sample Name: s98t002827_d Run Time: 12/01/98 14:52

Int. Std.	Ir[193]
Line	193/pulse
Units	.Cts/S
Avg	84.282
Stddev	.983
%RSD	1.1668

#1	85.612
#2	85.038
#3	83.803
#4	83.425
#5	83.532

Method: ACTINIDE Sample Name: s98t002827_a

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 14:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3391	.3980	274.0	2.472	.9226
Stddev	.0272	.0548	3.7	.138	.0266
%RSD	8.012	13.77	1.360	5.598	2.885

#1	.2982	.3459	270.5	2.588	.9384
#2	.3565	.3444	279.8	2.291	.9532
#3	.3698	.3934	271.4	2.494	.9088
#4	.3333	.4404	273.1	2.373	.8848
#5	.3378	.4659	275.1	2.613	.9279

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	561.5	687.6	17.99	.3960	.3532
Stddev	7.6	14.1	.20	.0290	.0358
%RSD	1.360	2.049	1.123	7.326	10.14

#1	560.4	671.9	17.70	.3579	.3579
#2	572.8	702.1	18.24	.4045	.3284
#3	564.1	690.7	17.91	.3738	.3108
#4	552.3	699.4	18.09	.4206	.3650
#5	557.9	673.8	18.02	.4232	.4038

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.680	.4000
Stddev	.093	.0279
%RSD	5.537	6.971

#1	1.746	.4175
#2	1.726	.3845
#3	1.755	.4406
#4	1.536	.3809
#5	1.638	.3766

Sample Name: s98t002827_a Run Time: 12/01/98 14:57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.330
Stddev	1.005
%RSD	1.1922

#1	83.833
#2	83.227
#3	84.724
#4	84.012
#5	85.854

Method: ACTINIDE Sample Name: s98t002940

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 15:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2962	.2960	.3959	1.849	.3000
Stddev	.0272	.0207	.0439	.071	.0316
%RSD	9.196	6.996	11.09	3.850	10.54

#1	.2518	.3091	.4426	1.889	.2556
#2	.2967	.3201	.3982	1.757	.3006
#3	.2973	.2934	.3320	1.934	.3436
#4	.3115	.2920	.3777	1.799	.2920
#5	.3238	.2653	.4292	1.869	.3082

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3309	.6490	.3377	.3659	.3271
Stddev	.0239	.0728	.0293	.0296	.0338
%RSD	7.226	11.22	8.664	8.081	10.32

#1	.3434	.7173	.3625	.3281	.3434
#2	.3201	.6636	.3084	.3982	.3591
#3	.3088	.7181	.3745	.3745	.2702
#4	.3154	.5723	.3154	.3426	.3309
#5	.3668	.5736	.3277	.3863	.3317

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3408	.3479
Stddev	.0333	.0146
%RSD	9.762	4.192

#1	.3663	.3701
#2	.2928	.3396
#3	.3590	.3397
#4	.3193	.3348
#5	.3668	.3551

Sample Name: s98t002940 Run Time: 12/01/98 15:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	86.029
Stdddev	.838
%RSD	.97466

#1	87.363
#2	85.392
#3	86.343
#4	85.615
#5	85.431

Method: ACTINIDE Sample Name: s98t002940_d

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 15:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3336	.3040	.4010	1.847	.3374
Stddev	.0138	.0334	.0473	.100	.0240
%RSD	4.146	10.98	11.78	5.430	7.105

#1	.3413	.3183	.3337	1.882	.3605
#2	.3435	.3435	.4284	1.783	.3087
#3	.3449	.2547	.4076	1.924	.3409
#4	.3247	.3131	.3788	1.704	.3169
#5	.3136	.2903	.4568	1.939	.3600

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3196	.5890	.3593	.3625	.3523
Stddev	.0311	.0497	.0423	.0313	.0345
%RSD	9.741	8.441	11.78	8.648	9.802

#1	.3260	.5791	.3337	.3797	.3605
#2	.2933	.6483	.3936	.3280	.3203
#3	.2900	.6231	.4036	.4076	.4076
#4	.3672	.5759	.3015	.3479	.3285
#5	.3213	.5187	.3639	.3484	.3445

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3353	.3283
Stddev	.0460	.0282
%RSD	13.72	8.585

#1	.3567	.3222
#2	.3010	.3280
#3	.4036	.3723
#4	.2899	.2937
#5	.3252	.3252

Sample Name: s98t002940_d Run Time: 12/01/98 15:13

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	86.140
Stddev	.677
%RSD	.78576

#1	86.913
#2	86.375
#3	85.059
#4	86.244
#5	86.107

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002940_a

Operator: kjt

Comment: 40501 df

Run Time: 12/01/98 15:18 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3342	.3678	291.9	3.767	.8830
Stddev	.0479	.0419	9.0	.171	.0421
%RSD	14.32	11.39	3.081	4.548	4.772
#1	.2563	.2991	288.0	3.601	.8856
#2	.3382	.3693	292.1	3.871	.9329
#3	.3304	.3765	287.0	3.658	.8299
#4	.3687	.3812	307.3	4.014	.9133
#5	.3776	.4130	285.1	3.689	.8535

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	587.4	726.8	19.21	.3790	.3412
Stddev	16.8	26.0	.64	.0520	.0223
%RSD	2.865	3.576	3.326	13.73	6.543
#1	578.9	698.7	18.34	.3884	.3340
#2	588.5	727.0	19.30	.3615	.3110
#3	583.6	714.2	19.11	.3420	.3420
#4	615.2	768.7	20.13	.4650	.3729
#5	570.8	725.4	19.15	.3383	.3461

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.746	.3360
Stddev	.131	.0246
%RSD	7.507	7.326
#1	1.674	.3418
#2	1.714	.3459
#3	1.625	.2997
#4	1.965	.3268
#5	1.750	.3658

Sample Name: s98t002940_a Run Time: 12/01/98 15:18

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.529
Stddev	2.865
%RSD	3.3897

#1	85.821
#2	85.756
#3	86.757
#4	79.564
#5	84.747

Method: ACTINIDE Sample Name: ccv

Operator: kjt

Comment: msact lcs

Run Time: 12/01/98 15:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2823	.3403	151.6	17.33	.6187
Stddev	.0358	.0369	2.1	.45	.0609
%RSD	12.69	10.84	1.370	2.616	9.836
#1	.2830	.3748	149.8	17.46	.7114
#2	.2421	.2844	152.4	17.20	.6419
#3	.2547	.3573	154.2	17.06	.5967
#4	.2996	.3220	149.2	16.88	.5916
#5	.3319	.3628	152.6	18.05	.5519

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	294.7	363.5	9.707	.3106	.3174
Stddev	4.8	6.1	.310	.0231	.0302
%RSD	1.638	1.675	3.196	7.446	9.517
#1	297.6	362.5	9.554	.2907	.3595
#2	293.4	360.1	9.985	.3228	.2844
#3	294.7	359.2	9.833	.3307	.3231
#4	287.5	361.3	9.237	.2808	.2921
#5	300.3	374.1	9.926	.3280	.3280

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.9697	.3309
Stddev	.0435	.0279
%RSD	4.488	8.447
#1	.9294	.3251
#2	.9801	.3267
#3	1.034	.3611
#4	.9286	.3520
#5	.9764	.2894

Sample Name: ccv Run Time: 12/01/98 15:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	87.394
Stddev	1.036
%RSD	1.1850

#1	87.150
#2	86.728
#3	87.700
#4	89.022
#5	86.372

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 12/01/98 15:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3088	.2749	.4013	.4197	.3034
Stddev	.0463	.0434	.0576	.0554	.0434
%RSD	15.00	15.80	14.36	13.21	14.31

#1	.3392	.3354	.4742	.3894	.3547
#2	.2883	.2691	.3191	.4959	.2768
#3	.2978	.2205	.4216	.4177	.2785
#4	.3690	.2537	.4152	.3498	.3460
#5	.2497	.2958	.3765	.4457	.2613

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2826	.6354	.3519	.3119	.3296
Stddev	.0412	.0392	.0109	.0277	.0425
%RSD	14.57	6.175	3.104	8.886	12.91

#1	.2814	.6245	.3431	.3354	.2891
#2	.2306	.6573	.3498	.2653	.3537
#3	.2591	.6498	.3404	.3094	.3017
#4	.3037	.6727	.3652	.3267	.3921
#5	.3381	.5725	.3612	.3227	.3112

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3550	.3504
Stddev	.0656	.0380
%RSD	18.48	10.86

#1	.3624	.2930
#2	.4536	.3344
#3	.3288	.3906
#4	.2729	.3613
#5	.3573	.3727

Sample Name: ccb Run Time: 12/01/98 15:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	86.566
Stddev	.243
%RSD	.28113

#1	86.466
#2	86.713
#3	86.182
#4	86.713
#5	86.756

LABCORE Completed Worklist Report for Worklist# 26709

Analyst: bjg

Instrument: ICPMS1

Book#: 990303-3 8859
11859

Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSACT TH232	QC	0.5e-2	4.96e-03	99.200 % Recovery	
1	ICV	0	QMSACT NP237	QC	4.94e-3	5.91e-03	119.636 % Recovery	
1	ICV	0	QMSACT PU239	QC	0.50e-2	5.21e-03	104.200 % Recovery	
1	ICV	0	QMSACT 241	QC	0.513e-2	5.63e-03	109.747 % Recovery	
2	ICB	0	QMSAC-QC PA231	QC	1	<5.57e-6	ug/mL	
2	ICB	0	QMSAC-QC TH229	QC	1	<4.46e-6	ug/mL	
2	ICB	0	QMSAC-QC TH230	QC	1	<3.55e-6	ug/mL	
2	ICB	0	QMSAC-QC TH232	QC	1	<1.93e-6	ug/mL	
2	ICB	0	QMSAC-QC NP237	QC	1	<4.67e-6	ug/mL	
2	ICB	0	QMSAC-QC PU239	QC	1	<3.47e-6	ug/mL	
2	ICB	0	QMSAC-QC PU240	QC	1	<3.55e-6	ug/mL	
2	ICB	0	QMSAC-QC PU242	QC	1	<2.78e-6	ug/mL	
2	ICB	0	QMSAC-QC 241	QC	1	<3.71e-6	ug/mL	
2	ICB	0	QMSAC-QC 243	QC	1	<3.83e-6	ug/mL	
2	ICB	0	QMSAC-QC 244	QC	1	<3.41e-6	ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 PA231-A1	LIQUID	N/A	< 2.256e-01	0.226 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 TH229-A1	LIQUID	N/A	< 1.807e-01	0.181 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 TH230-A1	LIQUID	N/A	< 1.438e-01	0.144 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 TH232-A1	LIQUID	N/A	< 7.817e-02	7.82e-002 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 NP237-A1	LIQUID	N/A	< 1.891e-01	0.189 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 PU239-A1	LIQUID	N/A	< 1.405e-01	0.141 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 PU240-A1	LIQUID	N/A	< 1.439e-01	0.144 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 PU242-A1	LIQUID	N/A	< 1.126e-01	0.113 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 241-A1	LIQUID	N/A	< 1.501e-01	0.150 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 243-A1	LIQUID	N/A	< 1.553e-01	0.155 ug/mL	
3	SAMPLE	S98T002826	0 A QMSACTA1 244-A1	LIQUID	N/A	< 1.383e-01	0.138 ug/mL	
4	DUP	S98T002826	0 A QMSACTA1 PA231-A1	LIQUID	<2.26e-1	< 2.26e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 TH229-A1	LIQUID	<1.81e-1	< 1.81e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 TH230-A1	LIQUID	<1.44e-1	< 1.44e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 TH232-A1	LIQUID	<7.82e-2	< 7.82e-2	RPD	
4	DUP	S98T002826	0 A QMSACTA1 NP237-A1	LIQUID	<1.89e-1	< 1.89e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 PU239-A1	LIQUID	<1.41e-1	< 1.41e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 PU240-A1	LIQUID	<1.44e-1	< 1.44e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 PU242-A1	LIQUID	<1.13e-1	< 1.13e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 241-A1	LIQUID	<1.50e-1	< 1.50e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 243-A1	LIQUID	<1.55e-1	< 1.55e-1	RPD	
4	DUP	S98T002826	0 A QMSACTA1 244-A1	LIQUID	<1.38e-1	< 1.38e-1	RPD	
5	SAMPLE	S98T002828	0 A QMSACTA1 PA231-A1	LIQUID	N/A	< 2.256e-01	0.226 ug/mL	
5	SAMPLE	S98T002828	0 A QMSACTA1 TH229-A1	LIQUID	N/A	< 1.807e-01	0.181 ug/mL	
5	SAMPLE	S98T002828	0 A QMSACTA1 TH230-A1	LIQUID	N/A	< 1.438e-01	0.144 ug/mL	
5	SAMPLE	S98T002828	0 A QMSACTA1 TH232-A1	LIQUID	N/A	< 7.817e-02	7.82e-002 ug/mL	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2074

LABCORE Completed Worklist Report for Worklist# 26709

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 SAMPLE	S98T002828	0 A	OMSACTA1 NF237-A1	LIQUID	N/A <	1.891e-01	0.189	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 PU239-A1	LIQUID	N/A <	1.405e-01	0.141	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 PU240-A1	LIQUID	N/A <	1.439e-01	0.144	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 PU242-A1	LIQUID	N/A <	1.126e-01	0.113	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 241-A1	LIQUID	N/A <	1.501e-01	0.150	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 243-A1	LIQUID	N/A <	1.553e-01	0.155	ug/mL
5 SAMPLE	S98T002828	0 A	OMSACTA1 244-A1	LIQUID	N/A <	1.383e-01	0.138	ug/mL
6 DUP	S98T002828	0 A	OMSACTA1 PA231-A1	LIQUID	<2.26e-1	<2.26e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 TH229-A1	LIQUID	<1.81e-1	<1.81e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 TH230-A1	LIQUID	<1.44e-1	<1.44e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 TH232-A1	LIQUID	<7.82e-2	<7.82e-2		RPD
6 DUP	S98T002828	0 A	OMSACTA1 NP237-A1	LIQUID	<1.89e-1	<1.89e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 PU239-A1	LIQUID	<1.41e-1	<1.41e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 PU240-A1	LIQUID	<1.44e-1	<1.44e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 PU242-A1	LIQUID	<1.13e-1	<1.13e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 241-A1	LIQUID	<1.50e-1	<1.50e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 243-A1	LIQUID	<1.55e-1	<1.55e-1		RPD
6 DUP	S98T002828	0 A	OMSACTA1 244-A1	LIQUID	<1.38e-1	<1.38e-1		RPD
7 SAMPLE	S98T002829	0 A	OMSACTA1 PA231-A1	LIQUID	N/A <	2.256e-01	0.226	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 TH229-A1	LIQUID	N/A <	1.807e-01	0.181	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 TH230-A1	LIQUID	N/A <	1.438e-01	0.144	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 TH232-A1	LIQUID	N/A <	7.817e-02	7.82e-002	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 NF237-A1	LIQUID	N/A <	1.891e-01	0.189	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 PU239-A1	LIQUID	N/A <	1.405e-01	0.141	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 PU240-A1	LIQUID	N/A <	1.439e-01	0.144	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 PU242-A1	LIQUID	N/A <	1.126e-01	0.113	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 241-A1	LIQUID	N/A <	1.501e-01	0.150	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 243-A1	LIQUID	N/A <	1.553e-01	0.155	ug/mL
7 SAMPLE	S98T002829	0 A	OMSACTA1 244-A1	LIQUID	N/A <	1.383e-01	0.138	ug/mL
8 DUP	S98T002829	0 A	OMSACTA1 PA231-A1	LIQUID	<2.26e-1	<2.26e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 TH229-A1	LIQUID	<1.81e-1	<1.81e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 TH230-A1	LIQUID	<1.44e-1	<1.44e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 TH232-A1	LIQUID	<7.82e-2	<7.82e-2		RPD
8 DUP	S98T002829	0 A	OMSACTA1 NP237-A1	LIQUID	<1.89e-1	<1.89e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 PU239-A1	LIQUID	<1.41e-1	<1.41e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 PU240-A1	LIQUID	<1.44e-1	<1.44e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 PU242-A1	LIQUID	<1.13e-1	<1.13e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 241-A1	LIQUID	<1.50e-1	<1.50e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 243-A1	LIQUID	<1.55e-1	<1.55e-1		RPD
8 DUP	S98T002829	0 A	OMSACTA1 244-A1	LIQUID	<1.38e-1	<1.38e-1		RPD
9 SAMPLE	S98T003385	0 D	OMSAC-D1 PA231-D1	LIQUID	N/A <	2.256e-01	0.226	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 TH229-D1	LIQUID	N/A <	1.807e-01	0.181	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 TH230-D1	LIQUID	N/A <	1.438e-01	0.144	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 TH232-D1	LIQUID	N/A <	7.817e-02	7.82e-002	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 NP237-D1	LIQUID	N/A <	1.891e-01	0.189	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 PU239-D1	LIQUID	N/A <	1.405e-01	0.141	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 PU240-D1	LIQUID	N/A <	1.439e-01	0.144	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 PU242-D1	LIQUID	N/A <	1.126e-01	0.113	ug/mL
9 SAMPLE	S98T003385	0 D	OMSAC-D1 241-D1	LIQUID	N/A <	1.501e-01	0.150	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26709

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 SAMPLE	S98T003385	0 D	@MSAC-D1 243-D1	LIQUID	N/A <	1.553e-01	0.155	ug/mL
9 SAMPLE	S98T003385	0 D	@MSAC-D1 244-D1	LIQUID	N/A <	1.383e-01	0.138	ug/mL
10 DUP	S98T003385	0 D	@MSAC-D1 P231-D1	LIQUID	<2.26e-1	<2.26e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 TH229-D1	LIQUID	<1.81e-1	<1.81e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 TH230-D1	LIQUID	<1.44e-1	<1.44e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 TH232-D1	LIQUID	<7.82e-2	<7.82e-2		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 NP237-D1	LIQUID	<1.89e-1	<1.89e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 PU239-D1	LIQUID	<1.41e-1	<1.41e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 PU240-D1	LIQUID	<1.44e-1	<1.44e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 PU242-D1	LIQUID	<1.13e-1	<1.13e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 241-D1	LIQUID	<1.50e-1	<1.50e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 243-D1	LIQUID	<1.55e-1	<1.55e-1		RPD
10 DUP	S98T003385	0 D	@MSAC-D1 244-D1	LIQUID	<1.38e-1	<1.38e-1		RPD
11 SPK-POST	S98T003385	0 D	@MSAC-D1 TH232-D1	LIQUID	0.020	1.88e-02	94.000 % Recovery	
11 SPK-POST	S98T003385	0 D	@MSAC-D1 NP237-D1	LIQUID	0.01976	2.29e-02	115.891 % Recovery	
11 SPK-POST	S98T003385	0 D	@MSAC-D1 PU239-D1	LIQUID	0.02008	1.99e-02	99.104 % Recovery	
11 SPK-POST	S98T003385	0 D	@MSAC-D1 241-D1	LIQUID	0.02052	2.16e-02	105.263 % Recovery	
12 CCV	0	0	@MSACT TH232	QC	0.5e-2	5.17e-03	103.400 % Recovery	
12 CCV	0	0	@MSACT NP237	QC	4.94e-3	6.05e-03	122.470 % Recovery	
12 CCV	0	0	@MSACT PU239	QC	0.50e-2	5.32e-03	106.400 % Recovery	
12 CCV	0	0	@MSACT 241	QC	0.513e-2	5.75e-03	112.086 % Recovery	
13 CCB	0	0	@MSAC-QC PA231	QC	1	<5.57e-6		ug/mL
13 CCB	0	0	@MSAC-QC TH229	QC	1	<4.46e-6		ug/mL
13 CCB	0	0	@MSAC-QC TH230	QC	1	<3.55e-6		ug/mL
13 CCB	0	0	@MSAC-QC TH232	QC	1	<1.93e-6		ug/mL
13 CCB	0	0	@MSAC-QC NP237	QC	1	<4.67e-6		ug/mL
13 CCB	0	0	@MSAC-QC PU239	QC	1	<3.47e-6		ug/mL
13 CCB	0	0	@MSAC-QC PU240	QC	1	<3.55e-6		ug/mL
13 CCB	0	0	@MSAC-QC PU242	QC	1	<2.78e-6		ug/mL
13 CCB	0	0	@MSAC-QC 241	QC	1	<3.71e-6		ug/mL
13 CCB	0	0	@MSAC-QC 243	QC	1	<3.83e-6		ug/mL
13 CCB	0	0	@MSAC-QC 244	QC	1	<3.41e-6		ug/mL

Final page for worklist# 26709

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

H. Quast
3-4-99
Np- 237 CCV recovery (122.5%) is noted. Samples will be rerun if requested.

LABCORE Data Entry Template for Worklist# 26709

Analyst: BSG Instrument: ICPMS1Book# 990303-3 8B59
11 B59Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICB			@MSACTA1	LIQUID		
3 SAMPLE	S98T002826 0 A		@MSACTA1	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
4 DUP	S98T002826 0 A		@MSACTA1	LIQUID		
5 SAMPLE	S98T002828 0 A		@MSACTA1	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
6 DUP	S98T002828 0 A		@MSACTA1	LIQUID		
7 SAMPLE	S98T002829 0 A		@MSACTA1	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
8 DUP	S98T002829 0 A		@MSACTA1	LIQUID		
9 SAMPLE	S98T003385 0 D		@MSAC-D1	LIQUID	98000437 BY-105 (2)	
	Analytes Requested: 241-D1 , 243-D1 , 244-D1 , NP237-D1, PA231-D1, PU239-D1, PU240-D1, PU242-D1, TH229-D1, TH230-D1, TH232-D1					
10 DUP	S98T003385 0 D		@MSAC-D1	LIQUID		
11 SPK-POST	S98T003385 0 D		@MSAC-D1	LIQUID		
12 CCV			@MSACT	QC		
13 CCB			@MSACTA1	LIQUID		

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 26709

S	Type	Sample#	R	A	Test	Matrix	Group#	Project

Final page for worklist # 26709

H. Anastas for BIG
Signature   Date   3-3-99

_____
Signature   Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/03/99 21:13 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	218400.
Stddev	1911.
%RSD	.8746

#1	222000.
#2	219800.
#3	218100.
#4	218400.
#5	219900.
#6	215500.
#7	218800.
#8	219000.
#9	216200.
#10	216800.

3-3-99

BY-MS

ACTIVATOR

SS8700 2826, 28, 29, 3385

.1-10-.025-00

FILE # ACO303A.725

W.C.F.S. 26208

27712

MS,ACT. 6ES DUTED 1:1 (770303.3)

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2079 TO 2107.

Method: ACTINIDE Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/03/99 21:14 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3554	.3488	.4708	.4100	.3619
Stddev	.0152	.0252	.0444	.0502	.0393
%RSD	4.269	7.218	9.439	12.25	10.87

#1	.3508	.3240	.4643	.4710	.3174
#2	.3570	.3639	.4748	.4367	.3223
#3	.3799	.3699	.3997	.4261	.4030
#4	.3504	.3673	.5020	.3572	.3875
#5	.3389	.3187	.5133	.3590	.3791

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3673	.5707	.3647	.3934	.3865
Stddev	.0359	.0473	.0477	.0328	.0499
%RSD	9.781	8.292	13.08	8.340	12.91

#1	.3541	.6514	.3675	.3975	.3374
#2	.3431	.5545	.3778	.4090	.4055
#3	.4096	.5648	.4261	.3369	.3865
#4	.4010	.5560	.2931	.4212	.3437
#5	.3288	.5267	.3590	.4026	.4596

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3663	.3534
Stddev	.0290	.0515
%RSD	7.905	14.57

#1	.3875	.3508
#2	.3639	.4402
#3	.3336	.3171
#4	.3437	.3471
#5	.4026	.3120

Analysis Report

HNF-1663 REV. 0

03/03/99 21:18:35

page 2

Sample Name: blank Run Time: 03/03/99 21:14

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	99.031
Stddev	1.760
%RSD	1.7774

#1	99.784
#2	96.174
#3	100.92
#4	98.928
#5	99.353

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 21:19 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3644	.3699	394.1	7.233	.8826
Stddev	.0234	.0358	9.0	.318	.0333
%RSD	6.431	9.683	2.285	4.402	3.774

#1	.3822	.3312	379.3	6.691	.9076
#2	.3915	.3685	392.2	7.445	.8389
#3	.3329	.4177	401.8	7.212	.8550
#4	.3521	.3916	397.4	7.359	.9116
#5	.3632	.3403	399.7	7.458	.8999

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	510.5	465.2	12.83	.4150	.3389
Stddev	11.3	13.5	.44	.0492	.0286
%RSD	2.215	2.905	3.432	11.86	8.452

#1	491.0	443.4	12.34	.4745	.3694
#2	513.9	467.9	12.45	.4507	.3553
#3	513.8	471.4	13.17	.3492	.2970
#4	513.2	479.5	13.35	.4081	.3488
#5	520.6	463.8	12.84	.3927	.3240

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	517.5	.3656
Stddev	11.1	.0206
%RSD	2.154	5.629

#1	498.1	.3885
#2	518.5	.3750
#3	522.1	.3524
#4	525.9	.3752
#5	522.7	.3371

Analysis Report

Sample Name: icv Run Time: 03/03/99 21:19

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.26
Stddev	1.40
%RSD	1.3654

#1	104.67
#2	101.32
#3	102.15
#4	101.29
#5	101.86

Analysis Report

Method: ACTINIDE Sample Name: ich

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 21:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3223	.3427	.4981	.3900	.3298
Stddev	.0291	.0311	.0663	.0491	.0092
%RSD	9.027	9.088	13.31	12.60	2.800

#1	.2770	.3682	.4595	.3480	.3209
#2	.3422	.3655	.5948	.3289	.3289
#3	.3473	.2910	.5060	.4432	.3208
#4	.3099	.3402	.4177	.4076	.3402
#5	.3350	.3484	.5125	.4221	.3383

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3312	.6645	.3548	.3580	.3521
Stddev	.0159	.0422	.0360	.0279	.0292
%RSD	4.800	6.347	10.16	7.803	8.290

#1	.3412	.6757	.3649	.3446	.3682
#2	.3389	.6678	.3356	.3190	.3157
#3	.3241	.6251	.3043	.3572	.3340
#4	.3065	.7276	.3705	.3840	.3907
#5	.3450	.6264	.3986	.3852	.3517

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3635	.3431
Stddev	.0390	.0474
%RSD	10.74	13.82

#1	.3987	.2770
#2	.2990	.3323
#3	.3572	.3572
#4	.3773	.3402
#5	.3852	.4087

Sample Name: icb Run Time: 03/03/99 21:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	99.648
Stddev	.896
%RSD	.89909

#1	98.666
#2	100.32
#3	100.79
#4	98.961
#5	99.503

Analysis Report

Method: ACTINIDE Sample Name: s98t002826

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:30 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3269	.3121	.4631	12.34	.3157
Stddev	.0322	.0406	.0435	.46	.0136
%RSD	9.863	13.01	9.400	3.740	4.293

#1	.3608	.3200	.4298	11.54	.3075
#2	.3389	.3522	.5017	12.42	.3024
#3	.2737	.2537	.5007	12.49	.3105
#4	.3330	.2902	.4781	12.67	.3363
#5	.3281	.3442	.4053	12.60	.3217

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3449	.5912	.3201	.3492	.3592
Stddev	.0385	.0605	.0434	.0298	.0270
%RSD	11.16	10.23	13.56	8.546	7.520

#1	.2792	.6588	.3577	.3294	.3294
#2	.3721	.5416	.2857	.4020	.3755
#3	.3605	.6376	.2904	.3372	.3972
#4	.3429	.6001	.2902	.3363	.3495
#5	.3699	.5179	.3764	.3410	.3442

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3437	.3353
Stddev	.0368	.0403
%RSD	10.70	12.02

#1	.3263	.3639
#2	.3289	.3588
#3	.3739	.3305
#4	.3001	.3561
#5	.3892	.2670

Sample Name: s98t002826 Run Time: 03/03/99 21:30

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.23
Stddev	2.68
%RSD	2.6190

#1	106.25
#2	100.32
#3	99.855
#4	101.09
#5	103.63

Method: ACTINIDE Sample Name: s98t002826_d

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:34 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2965	.3105	.5088	12.89	.3300
Stddev	.0112	.0159	.0370	.30	.0391
%RSD	3.773	5.132	7.276	2.355	11.84

#1	.2957	.3183	.5208	12.51	.3536
#2	.3137	.3073	.5337	12.68	.3849
#3	.2842	.2937	.5305	13.26	.3158
#4	.2898	.2994	.5152	13.10	.3091
#5	.2991	.3337	.4439	12.87	.2865

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3369	.5522	.3293	.2997	.3036
Stddev	.0270	.0502	.0479	.0146	.0187
%RSD	8.002	9.093	14.53	4.868	6.155

#1	.3247	.5915	.3665	.3118	.3279
#2	.3784	.6048	.3881	.2911	.3073
#3	.3063	.5653	.2747	.3000	.3095
#4	.3445	.4926	.2930	.3155	.2962
#5	.3306	.5069	.3243	.2802	.2771

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3504	.3049
Stddev	.0414	.0204
%RSD	11.81	6.678

#1	.3279	.3183
#2	.3978	.3267
#3	.2937	.3126
#4	.3799	.2866
#5	.3526	.2802

Sample Name: s98t002826_d Run Time: 03/03/99 21:34

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.34
Stddev	1.28
%RSD	1.2263

#1	103.69
#2	103.06
#3	105.56
#4	103.53
#5	105.87

Method: ACTINIDE Sample Name: s98t002828

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3091	.2854	.3717	1.054	.2856
Stddev	.0384	.0161	.0533	.064	.0395
%RSD	12.43	5.623	14.34	6.041	13.82

#1	.3454	.3084	.4502	1.101	.3238
#2	.3309	.2854	.3066	1.120	.2915
#3	.2885	.2915	.3462	.9566	.3097
#4	.3288	.2660	.3886	1.046	.2212
#5	.2517	.2759	.3669	1.046	.2820

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3067	.4801	.2946	.3060	.3111
Stddev	.0266	.0582	.0284	.0117	.0393
%RSD	8.684	12.13	9.625	3.812	12.63

#1	.3423	.4626	.3238	.3176	.3793
#2	.3249	.5465	.2884	.3006	.3036
#3	.2794	.4919	.2642	.2885	.2824
#4	.3019	.5082	.2720	.3139	.2870
#5	.2850	.3911	.3244	.3093	.3032

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3201	.2931
Stddev	.0542	.0119
%RSD	16.92	4.050

#1	.3947	.2868
#2	.3218	.2945
#3	.2490	.2763
#4	.3408	.3049
#5	.2941	.3032

Sample Name: s98t002828 Run Time: 03/03/99 21:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.82
Stddev	1.21
%RSD	1.1032

#1	108.09
#2	109.79
#3	109.77
#4	111.51
#5	109.94

Method: ACTINIDE Sample Name: s98t002828_d

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:42 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2984	.2999	.3563	1.044	.2766
Stddev	.0306	.0147	.0171	.069	.0255
%RSD	10.25	4.898	4.808	6.586	9.212
#1	.3046	.3046	.3314	1.054	.2359
#2	.2978	.2919	.3455	.9947	.2680
#3	.3444	.2895	.3657	1.158	.2865
#4	.2839	.2897	.3708	.9906	.2955
#5	.2613	.3236	.3682	1.021	.2969

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2812	.4217	.3141	.3001	.3050
Stddev	.0354	.0498	.0387	.0396	.0507
%RSD	12.60	11.82	12.31	13.19	16.61

#1	.2867	.3733	.3046	.2837	.3852
#2	.3216	.4676	.3782	.3395	.2859
#3	.3078	.4754	.3048	.3413	.3230
#4	.2433	.3708	.3099	.2868	.2665
#5	.2464	.4216	.2732	.2494	.2643

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2694	.3084
Stddev	.0337	.0241
%RSD	12.50	7.808

#1	.3076	.2956
#2	.2651	.3306
#3	.2621	.3383
#4	.2926	.2868
#5	.2197	.2910

Sample Name: s98t002828_d Run Time: 03/03/99 21:42

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	112.05
Stddev	2.03
%RSD	1.8155

#1	111.63
#2	111.92
#3	109.38
#4	115.08
#5	112.26

Method: ACTINIDE Sample Name: s98t002829

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:46 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3110	.2938	.3523	1.335	.3037
Stddev	.0272	.0540	.0571	.051	.0154
%RSD	8.757	18.37	16.21	3.806	5.087
#1	.3463	.2871	.2900	1.335	.3078
#2	.2934	.3776	.2934	1.275	.2934
#3	.3322	.2877	.3886	1.415	.2877
#4	.2813	.2900	.4146	1.334	.3277
#5	.3020	.2265	.3746	1.316	.3020

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2898	.4400	.3135	.3077	.2979
Stddev	.0304	.0502	.0289	.0252	.0242
%RSD	10.49	11.40	9.209	8.194	8.129
#1	.2723	.3551	.3019	.2752	.3078
#2	.2527	.4416	.3108	.3428	.3341
#3	.3322	.4716	.2729	.2966	.2877
#4	.3045	.4494	.3451	.3190	.2900
#5	.2875	.4821	.3369	.3049	.2701

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3176	.2791
Stddev	.0272	.0182
%RSD	8.552	6.515
#1	.2930	.2575
#2	.3341	.2818
#3	.2907	.2729
#4	.3161	.3074
#5	.3543	.2759

Sample Name: s98t002829 Run Time: 03/03/99 21:46

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	113.90
Stdddev	1.28
%RSD	1.1214

#1	112.63
#2	114.75
#3	112.38
#4	114.96
#5	114.78

Method: ACTINIDE Sample Name: s98t002829_d

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:50 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2784	.2824	.3502	1.327	.2936
Stddev	.0101	.0363	.0317	.074	.0242
%RSD	3.611	12.87	9.053	5.556	8.233
#1	.2799	.2621	.3454	1.453	.2561
#2	.2822	.2528	.3674	1.267	.3174
#3	.2674	.2556	.3438	1.308	.3115
#4	.2925	.3338	.3900	1.324	.2954
#5	.2701	.3075	.3046	1.284	.2873

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2947	.4218	.3093	.3215	.3153
Stddev	.0332	.0262	.0306	.0452	.0253
%RSD	11.26	6.222	9.890	14.06	8.037
#1	.2412	.4377	.3127	.3425	.3097
#2	.3086	.4027	.2910	.3615	.3498
#3	.3085	.4525	.3262	.3556	.2792
#4	.3279	.4284	.2688	.2895	.3161
#5	.2873	.3879	.3477	.2586	.3218

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3162	.3124
Stddev	.0401	.0236
%RSD	12.68	7.569
#1	.3544	.3276
#2	.2969	.3439
#3	.2733	.2938
#4	.3634	.2866
#5	.2931	.3103

Sample Name: s98t002829_d Run Time: 03/03/99 21:50

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	113.52
Stddev	1.51
%RSD	1.3337

#1	111.94
#2	113.41
#3	113.44
#4	112.83
#5	116.00

Method: ACTINIDE Sample Name: s98t003385

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:54 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2898	.2756	.3599	3.167	.2886
Stddev	.0391	.0459	.0191	.210	.0262
%RSD	13.51	16.64	5.296	6.646	9.094
#1	.2565	.3078	.3527	3.485	.3110
#2	.3243	.2201	.3778	2.960	.2459
#3	.3088	.3211	.3335	3.075	.2902
#4	.3201	.2960	.3564	3.047	.2839
#5	.2393	.2331	.3791	3.269	.3107

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3082	.4059	.2875	.2845	.2887
Stddev	.0304	.0470	.0180	.0274	.0227
%RSD	9.872	11.58	6.263	9.625	7.860
#1	.3367	.4649	.2822	.3236	.3142
#2	.2618	.4105	.2856	.2975	.3005
#3	.3304	.4230	.3119	.2717	.2532
#4	.3171	.3956	.2628	.2779	.2899
#5	.2952	.3356	.2952	.2517	.2859

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2945	.2885
Stddev	.0165	.0193
%RSD	5.596	6.696
#1	.2822	.2790
#2	.3153	.2975
#3	.2748	.3119
#4	.3050	.2930
#5	.2952	.2610

Sample Name: s98t003385   Run Time: 03/03/99 21:54

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	108.32
Stddev	3.09
%RSD	2.8570

#1	103.96
#2	112.04
#3	107.96
#4	110.37
#5	107.27

Method: ACTINIDE Sample Name: s98t003385_d

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 21:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2834	.2840	.3375	3.145	.2800
Stddev	.0511	.0306	.0337	.209	.0371
%RSD	18.05	10.78	9.996	6.638	13.24

#1	.2452	.2609	.3647	3.298	.2578
#2	.3313	.2425	.3372	3.067	.2840
#3	.3436	.3095	.3405	3.420	.2940
#4	.2305	.3102	.3634	3.029	.3309
#5	.2664	.2967	.2815	2.909	.2331

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2883	.3738	.2742	.3049	.3087
Stddev	.0146	.0253	.0428	.0163	.0284
%RSD	5.062	6.765	15.62	5.357	9.194

#1	.2955	.3993	.3364	.3112	.3395
#2	.2633	.3904	.2514	.3017	.2751
#3	.2879	.3560	.2321	.3281	.3033
#4	.2984	.3841	.2511	.2836	.2896
#5	.2967	.3391	.2997	.2997	.3360

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2912	.2778
Stddev	.0270	.0229
%RSD	9.279	8.254

#1	.3175	.2578
#2	.2662	.2840
#3	.3157	.2693
#4	.2600	.2630
#5	.2967	.3148

Sample Name: s98t003385_d Run Time: 03/03/99 21:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.87
Stddev	3.01
%RSD	2.7378

#1	106.03
#2	112.69
#3	107.69
#4	112.82
#5	110.11

Method: ACTINIDE Sample Name: s98t003385_a

Operator: bjg

Comment: 40501 df

Run Time: 03/03/99 22:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3009	.4432	1497.	6.636	2.381
Stddev	.0350	.0361	14.	.301	.127
%RSD	11.62	8.140	.9305	4.532	5.326

#1	.2527	.4653	1506.	7.004	2.496
#2	.2778	.4905	1479.	6.279	2.370
#3	.3128	.4160	1488.	6.884	2.247
#4	.3227	.4429	1501.	6.550	2.267
#5	.3386	.4015	1514.	6.463	2.523

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1973.	1775.	48.89	.5215	.2835
Stddev	30.	46.	1.20	.0569	.0252
%RSD	1.539	2.580	2.457	10.91	8.887

#1	2005.	1759.	49.11	.4930	.3205
#2	1923.	1751.	47.29	.4728	.2689
#3	1972.	1844.	50.14	.4767	.2976
#4	1980.	1725.	48.07	.5955	.2581
#5	1986.	1794.	49.86	.5693	.2726

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1982.	.3311
Stddev	55.	.0204
%RSD	2.773	6.156

#1	2032.	.3420
#2	1892.	.3368
#3	2015.	.2976
#4	1973.	.3285
#5	1997.	.3505

Sample Name: s98t003385_a Run Time: 03/03/99 22:02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.13
Stddev	2.22
%RSD	1.9969

#1	108.17
#2	112.81
#3	109.78
#4	113.64
#5	111.25

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 22:11 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3000	.3124	411.3	7.383	.8148
Stddev	.0253	.0286	5.2	.103	.0916
%RSD	8.422	9.162	1.265	1.392	11.24

#1	.3404	.2625	408.5	7.457	.8135
#2	.2856	.3341	407.9	7.391	.7254
#3	.3020	.3276	407.6	7.327	.7835
#4	.2734	.3190	412.6	7.240	.7832
#5	.2986	.3189	419.9	7.498	.9684

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	522.6	474.8	12.99	.3479	.3188
Stddev	6.8	5.9	.28	.0321	.0343
%RSD	1.300	1.247	2.135	9.234	10.77

#1	520.2	466.9	12.94	.3260	.2856
#2	516.0	481.3	13.13	.3941	.2856
#3	520.5	470.5	12.54	.3561	.3191
#4	522.2	478.4	13.25	.3532	.3646
#5	534.0	476.8	13.11	.3102	.3392

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	528.1	.3006
Stddev	4.8	.0505
%RSD	.9044	16.81

#1	533.1	.3318
#2	528.5	.2799
#3	523.8	.3476
#4	532.3	.2222
#5	522.7	.3218

Sample Name: ccv Run Time: 03/03/99 22:11

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	116.26
Stddev	.94
%RSD	.80934

#1	115.55
#2	116.72
#3	117.00
#4	117.04
#5	114.97

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 22:17 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2786	.2758	.5195	.3391	.2792
Stddev	.0110	.0287	.0533	.0245	.0339
%RSD	3.956	10.40	10.27	7.221	12.13

#1	.2692	.2352	.5610	.3174	.2947
#2	.2922	.2894	.5844	.3348	.2780
#3	.2829	.2915	.4515	.3172	.2315
#4	.2655	.3055	.4939	.3512	.2684
#5	.2833	.2576	.5065	.3749	.3234

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2861	.9542	.2981	.2713	.2968
Stddev	.0324	.0577	.0470	.0213	.0272
%RSD	11.33	6.044	15.78	7.835	9.155

#1	.2437	1.006	.2522	.2663	.3400
#2	.2780	1.021	.2667	.2582	.3036
#3	.2715	.9145	.3458	.3029	.2886
#4	.3198	.8879	.2741	.2484	.2826
#5	.3177	.9415	.3520	.2805	.2690

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2981	.2906
Stddev	.0339	.0265
%RSD	11.37	9.120

#1	.2975	.2833
#2	.2667	.2894
#3	.3515	.2972
#4	.2712	.3283
#5	.3034	.2547

Sample Name: ccb Run Time: 03/03/99 22:17

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	117.00
Stddev	.53
%RSD	.45280

#1	117.64
#2	117.50
#3	116.64
#4	116.76
#5	116.48

LABCORE Completed Worklist Report for Worklist# 26710

Analyst: bjg

Instrument: ICPMS1

Book#: 5859

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		EMSACT TH232	QC	1.0e-2	9.64e-03	96.400 % Recovery	
1 ICV	0		EMSACT NP237	QC	9.84e-3	1.36e-02	138.211 % Recovery	
1 ICV	0		EMSACT PU239	QC	1.05e-2	1.52e-02	144.762 % Recovery	
1 ICV	0		EMSACT 241	QC	n/a	2.20e-05	% Recovery	
2 ICB	0		EMSACT PA231-A1	LIQUID	1	<2.12e-5	ug/mL	
2 ICB	0		EMSACT PU239-A1	LIQUID	1	<1.85e-5	ug/mL	
2 ICB	0		EMSACT PU240-A1	LIQUID	1	<1.27e-5	ug/mL	
2 ICB	0		EMSACT PU242-A1	LIQUID	1	<7.34e-6	ug/mL	
2 ICB	0		EMSACT 241-A1	LIQUID	1	<9.45e-6	ug/mL	
2 ICB	0		EMSACT 243-A1	LIQUID	1	<1.36e-5	ug/mL	
2 ICB	0		EMSACT 244-A1	LIQUID	1	<1.15e-5	ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT PA231-A1	LIQUID	N/A	8.596e-01	0.859 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT TH229-A1	LIQUID	N/A	1.159e+00	1.159 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT TH230-A1	LIQUID	N/A	1.039e+00	1.039 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT TH232-A1	LIQUID	N/A	1.099e+00	1.099 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT NP237-A1	LIQUID	N/A	1.638e+00	1.638 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT PU239-A1	LIQUID	N/A	7.490e-01	0.749 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT PU240-A1	LIQUID	N/A	5.138e-01	0.514 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT PU242-A1	LIQUID	N/A	2.972e-01	0.297 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT 241-A1	LIQUID	N/A	3.828e-01	0.383 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT 243-A1	LIQUID	N/A	5.501e-01	0.550 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT 244-A1	LIQUID	N/A	4.638e-01	0.464 ug/mL	
4 DUP	S98T002968	0 A	EMSACT PA231-A1	LIQUID	<8.60e-1	<8.60e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT TH229-A1	LIQUID	<1.16e0	<1.16e0	RPD	
4 DUP	S98T002968	0 A	EMSACT TH230-A1	LIQUID	<1.04e0	<1.04e0	RPD	
4 DUP	S98T002968	0 A	EMSACT TH232-A1	LIQUID	<1.10e0	<1.10e0	RPD	
4 DUP	S98T002968	0 A	EMSACT NP237-A1	LIQUID	<1.64e0	<1.64e0	RPD	
4 DUP	S98T002968	0 A	EMSACT PU239-A1	LIQUID	<7.49e-1	<7.49e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT PU240-A1	LIQUID	<5.14e-1	<5.14e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT PU242-A1	LIQUID	<2.97e-1	<2.97e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT 241-A1	LIQUID	<3.83e-1	<3.83e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT 243-A1	LIQUID	<5.50e-1	<5.50e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT 244-A1	LIQUID	<4.64e-1	<4.64e-1	RPD	
5 SPK-POST	S98T002968	0 A	EMSACT TH232-A1	LIQUID	0.0210	1.80e-02	85.714 % Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT NP237-A1	LIQUID	0.0168	2.68e-02	159.524 % Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT PU239-A1	LIQUID	0.0210	3.05e-02	145.238 % Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT 241-A1	LIQUID	n/a	4.79e-05	% Recovery	
6 SAMPLE	S98T002969	0 A	EMSACT PA231-A1	LIQUID	N/A	8.596e-01	0.859 ug/mL	
6 SAMPLE	S98T002969	0 A	EMSACT TH229-A1	LIQUID	N/A	1.159e+00	1.159 ug/mL	
6 SAMPLE	S98T002969	0 A	EMSACT TH230-A1	LIQUID	N/A	1.039e+00	1.039 ug/mL	
6 SAMPLE	S98T002969	0 A	EMSACT TH232-A1	LIQUID	N/A	1.099e+00	1.099 ug/mL	

LABCORE Completed Worklist Report for Worklist# 26710

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002969	0 A	OMSACTAI	NP237-A1 LIQUID	N/A <	1.638e+00	1.638	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	FU239-A1 LIQUID	N/A <	7.490e-01	0.749	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	FU240-A1 LIQUID	N/A <	5.138e-01	0.514	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	FU242-A1 LIQUID	N/A <	2.972e-01	0.297	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	241-A1 LIQUID	N/A <	3.828e-01	0.383	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	243-A1 LIQUID	N/A <	5.501e-01	0.550	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAI	244-A1 LIQUID	N/A <	4.638e-01	0.464	ug/mL
7 DUP	S98T002969	0 A	OMSACTAI	PA231-A1 LIQUID	<8.60e-1	<8.60e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	TH229-A1 LIQUID	<1.16e0	<1.16e0		RPD
7 DUP	S98T002969	0 A	OMSACTAI	TH230-A1 LIQUID	<1.04e0	<1.04e0		RPD
7 DUP	S98T002969	0 A	OMSACTAI	TH232-A1 LIQUID	<1.10e0	<1.10e0		RPD
7 DUP	S98T002969	0 A	OMSACTAI	NP237-A1 LIQUID	<1.64e0	<1.64e0		RPD
7 DUP	S98T002969	0 A	OMSACTAI	FU239-A1 LIQUID	<7.49e-1	<7.49e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	FU240-A1 LIQUID	<5.14e-1	<5.14e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	FU242-A1 LIQUID	<2.97e-1	<2.97e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	241-A1 LIQUID	<3.83e-1	<3.83e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	243-A1 LIQUID	<5.50e-1	<5.50e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAI	244-A1 LIQUID	<4.64e-1	<4.64e-1		RPD
8 SAMPLE	S98T003010	0 A	OMSACTAI	PA231-A1 LIQUID	N/A <	8.596e-01	0.859	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	TH229-A1 LIQUID	N/A <	1.159e+00	1.159	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	TH230-A1 LIQUID	N/A <	1.039e+00	1.039	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	TH232-A1 LIQUID	N/A <	1.099e+00	1.099	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	NP237-A1 LIQUID	N/A <	1.638e+00	1.638	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	FU239-A1 LIQUID	N/A <	7.490e-01	0.749	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	FU240-A1 LIQUID	N/A <	5.138e-01	0.514	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	FU242-A1 LIQUID	N/A <	2.972e-01	0.297	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	241-A1 LIQUID	N/A <	3.828e-01	0.383	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	243-A1 LIQUID	N/A <	5.501e-01	0.550	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAI	244-A1 LIQUID	N/A <	4.638e-01	0.464	ug/mL
9 DUP	S98T003010	0 A	OMSACTAI	PA231-A1 LIQUID	<8.60e-1	<8.60e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	TH229-A1 LIQUID	<1.16e0	<1.16e0		RPD
9 DUP	S98T003010	0 A	OMSACTAI	TH230-A1 LIQUID	<1.04e0	<1.04e0		RPD
9 DUP	S98T003010	0 A	OMSACTAI	TH232-A1 LIQUID	<1.10e0	<1.10e0		RPD
9 DUP	S98T003010	0 A	OMSACTAI	NP237-A1 LIQUID	<1.64e0	<1.64e0		RPD
9 DUP	S98T003010	0 A	OMSACTAI	FU239-A1 LIQUID	<7.49e-1	<7.49e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	FU240-A1 LIQUID	<5.14e-1	<5.14e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	FU242-A1 LIQUID	<2.97e-1	<2.97e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	241-A1 LIQUID	<3.83e-1	<3.83e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	243-A1 LIQUID	<5.50e-1	<5.50e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAI	244-A1 LIQUID	<4.64e-1	<4.64e-1		RPD
10 SAMPLE	S98T003011	0 A	OMSACTAI	PA231-A1 LIQUID	N/A <	8.596e-01	0.859	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	TH229-A1 LIQUID	N/A <	1.159e+00	1.159	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	TH230-A1 LIQUID	N/A <	1.039e+00	1.039	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	TH232-A1 LIQUID	N/A <	1.099e+00	1.099	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	NP237-A1 LIQUID	N/A <	1.638e+00	1.638	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	FU239-A1 LIQUID	N/A <	7.490e-01	0.749	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	FU240-A1 LIQUID	N/A <	5.138e-01	0.514	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	FU242-A1 LIQUID	N/A <	2.972e-01	0.297	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAI	241-A1 LIQUID	N/A <	3.828e-01	0.383	ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26710

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
10 SAMPLE	S98T003011	0 A	EMSACTA1 243-A1	LIQUID	N/A <	5.501e-01	0.550	ug/mL
10 SAMPLE	S98T003011	0 A	EMSACTA1 244-A1	LIQUID	N/A <	4.638e-01	0.464	ug/mL
11 DUP	S98T003011	0 A	EMSACTA1 PA231-A1	LIQUID	<8.60e-1	<8.60e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 TH229-A1	LIQUID	<1.16e0	<1.16e0		RPD
11 DUP	S98T003011	0 A	EMSACTA1 TH230-A1	LIQUID	<1.04e0	<1.04e0		RPD
11 DUP	S98T003011	0 A	EMSACTA1 TH232-A1	LIQUID	<1.10e0	<1.10e0		RPD
11 DUP	S98T003011	0 A	EMSACTA1 NP237-A1	LIQUID	<1.64e0	<1.64e0		RPD
11 DUP	S98T003011	0 A	EMSACTA1 PU239-A1	LIQUID	<7.49e-1	<7.49e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 PU240-A1	LIQUID	<5.14e-1	<5.14e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 PU242-A1	LIQUID	<2.97e-1	<2.97e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 241-A1	LIQUID	<3.83e-1	<3.83e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 243-A1	LIQUID	<5.50e-1	<5.50e-1		RPD
11 DUP	S98T003011	0 A	EMSACTA1 244-A1	LIQUID	<4.64e-1	<4.64e-1		RPD
12 CCV		0	EMSACT TH232	QC	1.0e-2	8.54e-03	85.400 % Recovery	
12 CCV		0	EMSACT NP237	QC	9.84e-3	1.21e-02	122.967 % Recovery	
12 CCV		0	EMSACT PU239	QC	1.05e-2	1.42e-02	135.238 % Recovery	
12 CCV		0	EMSACT 241	QC	n/a	1.98e-05	% Recovery	
13 CCB		0	EMSACTA1 PA231-A1	LIQUID	1	<2.12e-5		ug/mL
13 CCB		0	EMSACTA1 TH229-A1	LIQUID	1	<2.86e-5		ug/mL
13 CCB		0	EMSACTA1 TH230-A1	LIQUID	1	<2.57e-5		ug/mL
13 CCB		0	EMSACTA1 TH232-A1	LIQUID	1	<2.71e-5		ug/mL
13 CCB		0	EMSACTA1 PU239-A1	LIQUID	1	<1.85e-5		ug/mL
13 CCB		0	EMSACTA1 PU240-A1	LIQUID	1	<1.27e-5		ug/mL
13 CCB		0	EMSACTA1 PU242-A1	LIQUID	1	<7.34e-6		ug/mL
13 CCB		0	EMSACTA1 241-A1	LIQUID	1	<9.45e-6		ug/mL
13 CCB		0	EMSACTA1 243-A1	LIQUID	1	<1.36e-5		ug/mL
13 CCB		0	EMSACTA1 244-A1	LIQUID	1	<1.15e-5		ug/mL

Final page for worklist# 26710

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sa M. Long 12/7/98
Reviewer Signature _____ Date _____

LABCORE Completed Worklist Report for Worklist# 26710

Analyst: bjg

Instrument: ICPMS1

Book# 5BS9

Method: LA-500-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		EMSACT TH232	QC	1.0e-2	9.64e-03	96.400 % Recovery	
1 ICV	0		EMSACT NP237	QC	9.84e-3	1.36e-02	138.211 % Recovery	
1 ICV	0		EMSACT PU239	QC	1.05e-2	1.52e-02	144.762 % Recovery	
1 ICV	0		EMSACT 241	QC	8.91e-3	2.20e-05	0.247 % Recovery	
2 ICB	0		EMSACT1 PA231-A1	LIQUID	1	<2.12e-5	ug/mL	
2 ICB	0		EMSACT1 TH229-A1	LIQUID	1	<2.86e-5	ug/mL	
2 ICB	0		EMSACT1 TH230-A1	LIQUID	1	<2.57e-5	ug/mL	
2 ICB	0		EMSACT1 TH232-A1	LIQUID	1	<2.71e-5	ug/mL	
2 ICB	0		EMSACT1 NP237-A1	LIQUID	1	<4.04e-5	ug/mL	
2 ICB	0		EMSACT1 PU239-A1	LIQUID	1	<1.85e-5	ug/mL	
2 ICB	0		EMSACT1 PU240-A1	LIQUID	1	<1.27e-5	ug/mL	
2 ICB	0		EMSACT1 PU242-A1	LIQUID	1	<7.34e-6	ug/mL	
2 ICB	0		EMSACT1 241-A1	LIQUID	1	<9.45e-6	ug/mL	
2 ICB	0		EMSACT1 243-A1	LIQUID	1	<1.36e-5	ug/mL	
2 ICB	0		EMSACT1 244-A1	LIQUID	1	<1.15e-5	ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 PA231-A1	LIQUID	N/A	< 8.596e-01	0.859 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 TH229-A1	LIQUID	N/A	< 1.159e+00	1.159 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 TH230-A1	LIQUID	N/A	< 1.039e+00	1.039 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 TH232-A1	LIQUID	N/A	< 1.099e+00	1.099 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 NP237-A1	LIQUID	N/A	< 1.638e+00	1.638 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 PU239-A1	LIQUID	N/A	< 7.490e-01	0.749 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 PU240-A1	LIQUID	N/A	< 5.138e-01	0.514 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 PU242-A1	LIQUID	N/A	< 2.972e-01	0.297 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 241-A1	LIQUID	N/A	< 3.828e-01	0.383 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 243-A1	LIQUID	N/A	< 5.501e-01	0.550 ug/mL	
3 SAMPLE	S98T002968	0 A	EMSACT1 244-A1	LIQUID	N/A	< 4.638e-01	0.464 ug/mL	
4 DUP	S98T002968	0 A	EMSACT1 PA231-A1	LIQUID	<8.60e-1	<8.60e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 TH229-A1	LIQUID	<1.16e0	<1.16e0	RPD	
4 DUP	S98T002968	0 A	EMSACT1 TH230-A1	LIQUID	<1.04e0	<1.04e0	RPD	
4 DUP	S98T002968	0 A	EMSACT1 TH232-A1	LIQUID	<1.10e0	<1.10e0	RPD	
4 DUP	S98T002968	0 A	EMSACT1 NP237-A1	LIQUID	<1.64e0	<1.64e0	RPD	
4 DUP	S98T002968	0 A	EMSACT1 PU239-A1	LIQUID	<7.49e-1	<7.49e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 PU240-A1	LIQUID	<5.14e-1	<5.14e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 PU242-A1	LIQUID	<2.97e-1	<2.97e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 241-A1	LIQUID	<3.83e-1	<3.83e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 243-A1	LIQUID	<5.50e-1	<5.50e-1	RPD	
4 DUP	S98T002968	0 A	EMSACT1 244-A1	LIQUID	<4.64e-1	<4.64e-1	RPD	
5 SPK-POST	S98T002968	0 A	EMSACT1 PA231-A1	LIQUID	n/a	3.28e-05	% Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT1 TH229-A1	LIQUID	n/a	5.99e-06	% Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT1 TH230-A1	LIQUID	n/a	-6.10e-6	% Recovery	
5 SPK-POST	S98T002968	0 A	EMSACT1 TH232-A1	LIQUID	2.10e-2	1.80e-02	85.714 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2111

LABCORE Completed Worklist Report for Worklist# 26710

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 SPK-POST	S98T002968	0 A	OMSACTAL NP237-A1	LIQUID	1.68e-2	2.68e-02	159.524	% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL PU239-A1	LIQUID	2.10e-2	3.05e-02	145.238	% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL PU240-A1	LIQUID	n/a	7.46e-04		% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL PU242-A1	LIQUID	n/a	-2.98e-6		% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL 241-A1	LIQUID	1.78e-2	4.79e-05	0.263	% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL 243-A1	LIQUID	n/a	-3.19e-6		% Recovery
5 SPK-POST	S98T002968	0 A	OMSACTAL 244-A1	LIQUID	?	-2.28e-6		% Recovery
6 SAMPLE	S98T002969	0 A	OMSACTAL PA231-A1	LIQUID	N/A	< 8.596e-01	0.859	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL TH229-A1	LIQUID	N/A	< 1.159e+00	1.159	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL TH230-A1	LIQUID	N/A	< 1.039e+00	1.039	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL TH232-A1	LIQUID	N/A	< 1.099e+00	1.099	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL NP237-A1	LIQUID	N/A	< 1.638e+00	1.638	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL PU239-A1	LIQUID	N/A	< 7.490e-01	0.749	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL PU240-A1	LIQUID	N/A	< 5.138e-01	0.514	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL PU242-A1	LIQUID	N/A	< 2.972e-01	0.297	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL 241-A1	LIQUID	N/A	< 3.828e-01	0.383	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL 243-A1	LIQUID	N/A	< 5.501e-01	0.550	ug/mL
6 SAMPLE	S98T002969	0 A	OMSACTAL 244-A1	LIQUID	N/A	< 4.638e-01	0.464	ug/mL
7 DUP	S98T002969	0 A	OMSACTAL PA231-A1	LIQUID	<8.60e-1	<8.60e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL TH229-A1	LIQUID	<1.16e0	<1.16e0		RPD
7 DUP	S98T002969	0 A	OMSACTAL TH230-A1	LIQUID	<1.04e0	<1.04e0		RPD
7 DUP	S98T002969	0 A	OMSACTAL TH232-A1	LIQUID	<1.10e0	<1.10e0		RPD
7 DUP	S98T002969	0 A	OMSACTAL NP237-A1	LIQUID	<1.64e0	<1.64e0		RPD
7 DUP	S98T002969	0 A	OMSACTAL PU239-A1	LIQUID	<7.49e-1	<7.49e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL PU240-A1	LIQUID	<5.14e-1	<5.14e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL PU242-A1	LIQUID	<2.97e-1	<2.97e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL 241-A1	LIQUID	<3.83e-1	<3.83e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL 243-A1	LIQUID	<5.50e-1	<5.50e-1		RPD
7 DUP	S98T002969	0 A	OMSACTAL 244-A1	LIQUID	<4.64e-1	<4.64e-1		RPD
8 SAMPLE	S98T003010	0 A	OMSACTAL PA231-A1	LIQUID	N/A	< 8.596e-01	0.859	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL TH229-A1	LIQUID	N/A	< 1.159e+00	1.159	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL TH230-A1	LIQUID	N/A	< 1.039e+00	1.039	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL TH232-A1	LIQUID	N/A	< 1.099e+00	1.099	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL NP237-A1	LIQUID	N/A	< 1.638e+00	1.638	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL PU239-A1	LIQUID	N/A	< 7.490e-01	0.749	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL PU240-A1	LIQUID	N/A	< 5.138e-01	0.514	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL PU242-A1	LIQUID	N/A	< 2.972e-01	0.297	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL 241-A1	LIQUID	N/A	< 3.828e-01	0.383	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL 243-A1	LIQUID	N/A	< 5.501e-01	0.550	ug/mL
8 SAMPLE	S98T003010	0 A	OMSACTAL 244-A1	LIQUID	N/A	< 4.638e-01	0.464	ug/mL
9 DUP	S98T003010	0 A	OMSACTAL PA231-A1	LIQUID	<8.60e-1	<8.60e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAL TH229-A1	LIQUID	<1.16e0	<1.16e0		RPD
9 DUP	S98T003010	0 A	OMSACTAL TH230-A1	LIQUID	<1.04e0	<1.04e0		RPD
9 DUP	S98T003010	0 A	OMSACTAL TH232-A1	LIQUID	<1.10e0	<1.10e0		RPD
9 DUP	S98T003010	0 A	OMSACTAL NP237-A1	LIQUID	<1.64e0	<1.64e0		RPD
9 DUP	S98T003010	0 A	OMSACTAL PU239-A1	LIQUID	<7.49e-1	<7.49e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAL PU240-A1	LIQUID	<5.14e-1	<5.14e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAL PU242-A1	LIQUID	<2.97e-1	<2.97e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAL 241-A1	LIQUID	<3.83e-1	<3.83e-1		RPD

LABCORE Completed Worklist Report for Worklist# 26710

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 DUP	S98T003010	0 A	OMSACTAL 243-A1	LIQUID	<5.50e-1	<5.50e-1		RPD
9 DUP	S98T003010	0 A	OMSACTAL 244-A1	LIQUID	<4.64e-1	<4.64e-1		RPD
10 SAMPLE	S98T003011	0 A	OMSACTAL PA231-A1	LIQUID	N/A	8.596e-01	0.859	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL TH229-A1	LIQUID	N/A	1.159e+00	1.159	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL TH230-A1	LIQUID	N/A	1.039e+00	1.039	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL TH232-A1	LIQUID	N/A	1.099e+00	1.099	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL NP237-A1	LIQUID	N/A	1.638e+00	1.638	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL PU239-A1	LIQUID	N/A	7.490e-01	0.749	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL PU240-A1	LIQUID	N/A	5.138e-01	0.514	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL PU242-A1	LIQUID	N/A	2.972e-01	0.297	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL 241-A1	LIQUID	N/A	3.828e-01	0.383	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL 243-A1	LIQUID	N/A	5.501e-01	0.550	ug/mL
10 SAMPLE	S98T003011	0 A	OMSACTAL 244-A1	LIQUID	N/A	4.638e-01	0.464	ug/mL
11 DUP	S98T003011	0 A	OMSACTAL PA231-A1	LIQUID	<8.60e-1	<8.60e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL TH229-A1	LIQUID	<1.16e0	<1.16e0		RPD
11 DUP	S98T003011	0 A	OMSACTAL TH230-A1	LIQUID	<1.04e0	<1.04e0		RPD
11 DUP	S98T003011	0 A	OMSACTAL TH232-A1	LIQUID	<1.10e0	<1.10e0		RPD
11 DUP	S98T003011	0 A	OMSACTAL NP237-A1	LIQUID	<1.64e0	<1.64e0		RPD
11 DUP	S98T003011	0 A	OMSACTAL PU239-A1	LIQUID	<7.49e-1	<7.49e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL PU240-A1	LIQUID	<5.14e-1	<5.14e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL PU242-A1	LIQUID	<2.97e-1	<2.97e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL 241-A1	LIQUID	<3.83e-1	<3.83e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL 243-A1	LIQUID	<5.50e-1	<5.50e-1		RPD
11 DUP	S98T003011	0 A	OMSACTAL 244-A1	LIQUID	<4.64e-1	<4.64e-1		RPD
12 CCV	0	0	OMSACT TH232	QC	1.0e-2	8.54e-03	85.400 % Recovery	
12 CCV	0	0	OMSACT NP237	QC	9.84e-3	1.21e-02	122.967 % Recovery	
12 CCV	0	0	OMSACT PU239	QC	1.05e-2	1.42e-02	135.238 % Recovery	
12 CCV	0	0	OMSACT 241	QC	8.91e-3	1.98e-05	0.222 % Recovery	
13 CCB	0	0	OMSACTAL PA231-A1	LIQUID	1	<2.12e-5		ug/mL
13 CCB	0	0	OMSACTAL TH229-A1	LIQUID	1	<2.86e-5		ug/mL
13 CCB	0	0	OMSACTAL TH230-A1	LIQUID	1	<2.57e-5		ug/mL
13 CCB	0	0	OMSACTAL TH232-A1	LIQUID	1	<2.71e-5		ug/mL
13 CCB	0	0	OMSACTAL NP237-A1	LIQUID	1	<4.04e-5		ug/mL
13 CCB	0	0	OMSACTAL PU239-A1	LIQUID	1	<1.85e-5		ug/mL
13 CCB	0	0	OMSACTAL PU240-A1	LIQUID	1	<1.27e-5		ug/mL
13 CCB	0	0	OMSACTAL PU242-A1	LIQUID	1	<7.34e-6		ug/mL
13 CCB	0	0	OMSACTAL 241-A1	LIQUID	1	<9.45e-6		ug/mL
13 CCB	0	0	OMSACTAL 243-A1	LIQUID	1	<1.36e-5		ug/mL
13 CCB	0	0	OMSACTAL 244-A1	LIQUID	1	<1.15e-5		ug/mL

Final page for worklist# 26710

for: *Sal H. Pang*

BJG 12/2/98
_____
Analyst Signature Date

_____
Analyst Signature Date

Sal H. Pang 12/7/98
_____
Reviewer Signature Date

LABCORE Completed Worklist Report for Worklist# 26710

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

10/22/98 12:22
A-0004-1

FILE # AC1202A.TXT

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 26710

Analyst: B. GOECKE Instrument: ICPMS1 Book# 51559

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICE			@MSACTA1	LIQUID		
3 SAMPLE	S98T002968 0 A		@MSACTA1	LIQUID	98000509 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
4 DUP	S98T002968 0 A		@MSACTA1	LIQUID		
5 SPK - Post 9/24/98	S98T002968 0 A		@MSACTA1	LIQUID		
6 SAMPLE	S98T002969 0 A		@MSACTA1	LIQUID	98000509 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
7 DUP	S98T002969 0 A		@MSACTA1	LIQUID		
8 SAMPLE	S98T003010 0 A		@MSACTA1	LIQUID	98000510 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
9 DUP	S98T003010 0 A		@MSACTA1	LIQUID		
10 SAMPLE	S98T003011 0 A		@MSACTA1	LIQUID	98000510 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
11 DUP	S98T003011 0 A		@MSACTA1	LIQUID		
12 CCV			@MSACT	QC		
13 CCB			@MSACTA1	LIQUID		
14 SAMPLE	~~S98T003012 0 A~~		@MSACTA1	LIQUID	98000510 BY-105 (2)	
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					

Data Entry Comments:

Updated & Validated by:

Shail M. Puri

12/2/98

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2115

10/22/98 12:22
A-0004-1**LABCORE Data Entry Template for Worklist# 26710**

S Type	Sample#	R A	Test	Matrix	Group#	Project
15- ~~DUP~~ <i>off</i> <i>12-2-98</i>	~~598T003012~~ <i>0-A</i>		@MSACTA1	LIQUID		
16- ~~SPK~~ <i>off</i> <i>12-2-98</i>	~~598T003012~~ <i>0-A</i>		@MSACTA1	LIQUID		
17- ~~SAMPLE~~ <i>off</i> <i>12-2-98</i>	~~598T003013~~ <i>0-A</i>		@MSACTA1	LIQUID	98000510	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
18- ~~DUP~~ <i>off</i> <i>12-2-98</i>	~~598T003013~~ <i>0-A</i>		@MSACTA1	LIQUID		
19- ~~CCV~~ <i>off</i> <i>12-2-98</i>			@MSACT	QC		
20- ~~CCB~~ <i>off</i> <i>12-2-98</i>			@MSACTA1	LIQUID		

Final page for worklist # 26710

[Signature] *12-2-98*
Analyst Signature Date

_____
Analyst Signature Date

598T002968, 69, 3010, 11, 12, 13 *off 12-2-98*

1-10-1 - .025-10-1 *D.F.*
40501

* SPK = .04ml of MSACT SPK 41859

SCANS = 98002968
69
3010
11
12 off
13 12-2-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

HNF-1663 REV. 0

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 12/02/98 12:56 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	144600.
Stddev	892.
%RSD	.6167

#1	144400.
#2	144700.
#3	143700.
#4	143700.
#5	144700.
#6	146100.
#7	143900.
#8	144500.
#9	146200.
#10	144000.

*Effect**12-2-98**BY-105**AF**1-10-1-1025-10-1 40301**588T00 2968, 67, 3010, 11, 12, 13**FILE # AL1202A.TXT*

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2117 TO 2119.

Method: ACTINIDE Sample Name: cal blank

Operator: djg

Comment: background

Run Time: 12/02/98 12:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4512	.4933	.5526	.5163	.4608
Stddev	.0467	.0465	.0590	.0197	.0423
%RSD	10.34	9.433	10.67	3.819	9.191

#1	.4683	.5176	.5669	.4847	.4642
#2	.4336	.4749	.5368	.5286	.4666
#3	.3991	.4197	.6377	.5349	.4073
#4	.5223	.5307	.5476	.5223	.4423
#5	.4328	.5235	.4740	.5111	.5235

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5184	.5930	.5297	.5312	.4873
Stddev	.0871	.0445	.0459	.0292	.0497
%RSD	16.80	7.511	8.662	5.492	10.20

#1	.5012	.5628	.4683	.5299	.4272
#2	.4708	.6277	.5534	.5451	.5575
#3	.6706	.6336	.4937	.4855	.5102
#4	.4549	.5307	.5602	.5307	.4591
#5	.4946	.6100	.5729	.5647	.4822

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5445	.5076
Stddev	.0359	.0565
%RSD	6.597	11.13

#1	.5176	.5669
#2	.5905	.4708
#3	.5431	.5472
#4	.5686	.4296
#5	.5028	.5235

Analysis Report

HNF-1663 REV. 0

12/02/98 13:02:30

page 2

Sample Name: cal blank Run Time: 12/02/98 12:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.577
Stddev	.821
%RSD	1.0183

#1	81.143
#2	80.718
#3	81.016
#4	79.137
#5	80.873

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.010 ppm cal

Run Time: 12/02/98 13:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4493	.4517	152.7	228.5	.7859
Stddev	.0539	.0615	3.8	3.2	.0215
%RSD	11.99	13.62	2.463	1.393	2.735

#1	.5280	.3844	149.6	226.6	.8152
#2	.4542	.5375	148.0	228.7	.7709
#3	.4662	.4195	153.9	233.2	.7755
#4	.3928	.4910	156.7	229.0	.8020
#5	.4055	.4260	155.5	224.7	.7659

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.004	.7304	.4860	.5060	.5047
Stddev	.023	.0522	.0645	.0622	.0331
%RSD	2.254	7.145	13.27	12.29	6.562

#1	.9841	.6589	.5406	.5322	.4942
#2	.9792	.7709	.5542	.4792	.5125
#3	1.021	.6992	.4958	.5891	.5424
#4	1.031	.7365	.4133	.4215	.4542
#5	1.003	.7864	.4260	.5073	.5202

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5079	.4977
Stddev	.0634	.0695
%RSD	12.49	13.96

#1	.5533	.5575
#2	.4125	.4292
#3	.5255	.5806
#4	.4787	.4869
#5	.5693	.4342

Sample Name: cal std Run Time: 12/02/98 13:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.084
Stddev	1.322
%RSD	1.6510

#1	78.921
#2	79.998
#3	78.656
#4	81.464
#5	81.384

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 12/02/98 13:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4540	.6517	1537.	2252.	3.897
Stddev	.0578	.0595	14.	39.	.125
%RSD	12.73	9.124	.9173	1.714	3.215

#1	.4584	.7062	1555.	2293.	3.948
#2	.4170	.5957	1549.	2269.	3.717
#3	.4348	.7247	1526.	2263.	3.958
#4	.4081	.6244	1528.	2190.	4.036
#5	.5515	.6074	1525.	2247.	3.828

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.041	1.302	.4589	.4922	.4719
Stddev	.248	.094	.0238	.0331	.0081
%RSD	4.104	7.243	5.188	6.723	1.720

#1	6.343	1.404	.4667	.5286	.4832
#2	5.730	1.386	.4765	.4964	.4765
#3	5.878	1.180	.4831	.5194	.4710
#4	6.036	1.294	.4408	.4571	.4653
#5	6.218	1.247	.4276	.4596	.4636

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4679	.4798
Stddev	.0546	.0532
%RSD	11.67	11.09

#1	.5204	.5575
#2	.5242	.4368
#3	.3946	.4952
#4	.4530	.4857
#5	.4476	.4236

Sample Name: cal std Run Time: 12/02/98 13:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	82.507
Stddev	1.310
%RSD	1.5873

#1	80.713
#2	83.938
#3	82.794
#4	81.676
#5	83.414

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 12/02/98 13:19 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3934	.3869	147.2	16.95	.6758
Stddev	.0270	.0328	3.3	.35	.0620
%RSD	6.875	8.482	2.263	2.090	9.179
#1	.4246	.3634	144.1	16.49	.5929
#2	.3853	.3853	144.4	17.31	.6499
#3	.4160	.3544	145.8	16.75	.6664
#4	.3578	.3932	150.7	17.29	.7156
#5	.3832	.4385	150.8	16.89	.7545

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	292.8	366.8	9.856	.4469	.4302
Stddev	4.1	8.7	.175	.0734	.0110
%RSD	1.389	2.364	1.775	16.43	2.565
#1	290.3	355.2	9.685	.3251	.4169
#2	293.6	367.9	9.940	.4670	.4320
#3	288.2	361.5	9.715	.5123	.4468
#4	299.0	372.7	10.11	.4915	.4246
#5	292.7	376.8	9.828	.4385	.4306

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.129	.4429
Stddev	.108	.0500
%RSD	9.573	11.29
#1	.9984	.4016
#2	1.027	.3775
#3	1.233	.4777
#4	1.203	.4679
#5	1.185	.4898

Sample Name: icv Run Time: 12/02/98 13:19

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	85.699
Stddev	1.155
%RSD	1.3479

#1	87.141
#2	85.651
#3	86.534
#4	84.783
#5	84.388

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 12/02/98 13:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4167	.4172	.5531	.4644	.4217
Stddev	.0309	.0111	.0452	.0550	.0654
%RSD	7.420	2.664	8.172	11.84	15.50
#1	.3907	.4099	.5210	.5057	.3295
#2	.4458	.4302	.5123	.3989	.4224
#3	.4519	.4188	.5473	.4312	.4976
#4	.4104	.4025	.5579	.5340	.4662
#5	.3848	.4244	.6267	.4522	.3927

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4311	.6614	.4291	.4318	.4465
Stddev	.0574	.0490	.0508	.0358	.0388
%RSD	13.31	7.409	11.85	8.288	8.689
#1	.4405	.6321	.3678	.4329	.4291
#2	.3442	.6023	.4224	.4263	.4693
#3	.4685	.7007	.4644	.3815	.4395
#4	.4104	.7213	.4941	.4822	.4981
#5	.4919	.6505	.3967	.4363	.3967

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4488	.4575
Stddev	.0770	.0784
%RSD	17.17	17.14
#1	.3678	.3218
#2	.3637	.4615
#3	.5224	.5141
#4	.5061	.4862
#5	.4839	.5038

Sample Name: icb Run Time: 12/02/98 13:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.064
Stddev	2.435
%RSD	2.8965

#1	87.014
#2	85.231
#3	80.393
#4	83.648
#5	84.034

Method: ACTINIDE Sample Name: s98t002968

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 13:32 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4011	.3400	.5089	.7912	.3852
Stddev	.0542	.0304	.0372	.0682	.0375
%RSD	13.52	8.938	7.318	8.625	9.730
#1	.3542	.3621	.5195	.7675	.4330
#2	.4133	.3378	.5404	.7629	.3695
#3	.3422	.3462	.5451	.9073	.3462
#4	.4758	.2895	.4758	.7296	.4164
#5	.4202	.3647	.4638	.7888	.3607

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3995	.5786	.4154	.4257	.4456
Stddev	.0527	.0449	.0522	.0463	.0351
%RSD	13.19	7.758	12.57	10.87	7.873
#1	.4054	.5983	.4093	.4330	.3936
#2	.3735	.6318	.4967	.3695	.4411
#3	.3780	.5412	.4258	.4258	.4377
#4	.3529	.5234	.3886	.4957	.4719
#5	.4876	.5986	.3568	.4043	.4836

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4201	.4049
Stddev	.0396	.0668
%RSD	9.418	16.50
#1	.4448	.5117
#2	.3854	.4172
#3	.3780	.3422
#4	.4719	.3569
#5	.4202	.3964

Analysis Report

HNF-1663 REV. 0

12/02/98 13:36:13

page 2

Sample Name: s98t002968 Run Time: 12/02/98 13:32

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.100
Stddev	.355
%RSD	.42212
#1	84.690
#2	83.887
#3	83.769
#4	84.063
#5	84.091

Method: ACTINIDE Sample Name: s98t002968_d

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 13:38 Type: Unk

Mode: IR

Corr.Fact.: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3876	.3778	.4935	.7782	.4022
Stddev	.0710	.0393	.0414	.0493	.0372
%RSD	18.31	10.40	8.383	6.333	9.246

#1	.3482	.3834	.4733	.8567	.4186
#2	.4913	.3866	.4953	.7933	.3463
#3	.4316	.4076	.5155	.7513	.4476
#4	.3377	.3099	.4370	.7310	.4053
#5	.3292	.4015	.5461	.7589	.3935

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4676	.5291	.4310	.4301	.4077
Stddev	.0355	.0439	.0255	.0525	.0338
%RSD	7.597	8.304	5.921	12.21	8.296

#1	.4616	.5164	.4342	.4303	.4616
#2	.4349	.5718	.3947	.3423	.3745
#3	.4835	.5075	.4516	.4755	.3956
#4	.5205	.5761	.4569	.4643	.4172
#5	.4377	.4738	.4176	.4377	.3895

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4663	.4079
Stddev	.0669	.0355
%RSD	14.34	8.711

#1	.4342	.4108
#2	.3826	.4027
#3	.5315	.4635
#4	.4450	.3973
#5	.5380	.3654

Sample Name: s98t002968_d Run Time: 12/02/98 13:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	83.663
Stddev	.964
%RSD	1.1516

#1	85.209
#2	82.773
#3	83.416
#4	83.899
#5	83.018

Method: ACTINIDE Sample Name: s98t002968_a

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 13:43 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3828	.4159	274.6	2.897	.9181
Stddev	.0584	.0480	9.4	.075	.0744
%RSD	15.25	11.55	3.429	2.631	8.100

#1	.4126	.4164	261.8	2.834	.9588
#2	.3367	.3488	277.0	2.924	.8275
#3	.3438	.4397	274.5	2.910	.8475
#4	.4720	.3969	287.9	3.004	.9775
#5	.3491	.4775	271.6	2.813	.9791

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	578.2	735.5	19.41	.4483	.4180
Stddev	17.0	31.1	.75	.0626	.0621
%RSD	2.944	4.231	3.868	13.97	14.86

#1	557.8	697.0	18.36	.3895	.3705
#2	574.5	751.8	20.05	.4259	.4989
#3	571.5	725.8	18.92	.4037	.3518
#4	603.9	779.0	20.06	.4804	.4637
#5	583.4	723.9	19.65	.5417	.4053

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.818	.4148
Stddev	.103	.0426
%RSD	5.680	10.27

#1	1.646	.3744
#2	1.797	.4178
#3	1.887	.4037
#4	1.888	.3927
#5	1.874	.4856

Sample Name: s98t002968_a . Run Time: 12/02/98 13:43

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	83.137
Stddev	2.699
%RSD	3.2466

#1	87.259
#2	82.180
#3	83.387
#4	79.795
#5	83.067

Method: ACTINIDE Sample Name: s98t002969

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 13:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa [231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3821	.4034	.5118	.6795	.3974
Stddev	.0569	.0583	.1063	.0547	.0215
%RSD	14.88	14.45	20.77	8.045	5.401

#1	.3424	.4902	.5991	.7586	.3929
#2	.4621	.3611	.5514	.6640	.3767
#3	.4223	.4373	.3996	.6522	.4147
#4	.3343	.3606	.6123	.7062	.4245
#5	.3496	.3676	.3964	.6163	.3784

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4045	.8649	.4389	.3966	.4667
Stddev	.0762	.0868	.0414	.0516	.0428
%RSD	18.84	10.04	9.431	13.01	9.173

#1	.4552	.9415	.4902	.3963	.4902
#2	.3534	.9397	.4543	.3844	.4155
#3	.5090	.8935	.3846	.4713	.5203
#4	.3230	.8039	.4545	.3263	.4320
#5	.3820	.7460	.4108	.4035	.4757

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4749	.4394
Stddev	.0282	.0441
%RSD	5.944	10.03

#1	.4513	.4007
#2	.4543	.4077
#3	.5203	.5090
#4	.4658	.4545
#5	.4829	.4253

Sample Name: s98t002969 Run Time: 12/02/98 13:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	88.234
Stddev	2.769
%RSD	3.1379

#1	85.681
#2	85.842
#3	88.414
#4	88.739
#5	92.492

Method: ACTINIDE Sample Name: s98t002969_d

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 13:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa [231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3688	.3802	.4564	.6656	.4211
Stddev	.0604	.0281	.0555	.0521	.0341
%RSD	16.37	7.381	12.17	7.823	8.091

#1	.3110	.4025	.4574	.7025	.4464
#2	.3494	.3978	.5204	.7323	.4386
#3	.4691	.3972	.3707	.6318	.4388
#4	.3737	.3662	.4484	.6577	.4185
#5	.3408	.3371	.4852	.6038	.3630

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4250	.7153	.3844	.4017	.4359
Stddev	.0344	.0916	.0429	.0201	.0224
%RSD	8.098	12.80	11.15	5.014	5.148

#1	.3769	.7062	.4537	.3983	.4245
#2	.4721	.5687	.3680	.4312	.4572
#3	.4350	.8133	.3783	.3821	.4161
#4	.4260	.7250	.3849	.4111	.4185
#5	.4149	.7631	.3371	.3852	.4630

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3994	.4315
Stddev	.0337	.0214
%RSD	8.450	4.957

#1	.3915	.4208
#2	.4498	.4498
#3	.3669	.4502
#4	.3737	.3999
#5	.4149	.4371

Sample Name: s98t002969_d Run Time: 12/02/98 13:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	89.614
Stddev	1.091
%RSD	1.2177

#1	91.097
#2	89.671
#3	88.115
#4	89.200
#5	89.989

Method: ACTINIDE Sample Name: s98t003010

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 14:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3716	.3662	.4741	.7320	.3916
Stddev	.0275	.0337	.0518	.0676	.0336
%RSD	7.388	9.197	10.94	9.232	8.583

#1	.3495	.4035	.4900	.6125	.3603
#2	.3975	.3864	.5127	.7765	.4421
#3	.4050	.3571	.3829	.7658	.4087
#4	.3481	.3148	.4888	.7554	.3777
#5	.3580	.3692	.4960	.7496	.3692

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4151	.6461	.4039	.4052	.4492
Stddev	.0655	.0493	.0733	.0559	.0292
%RSD	15.77	7.624	18.14	13.79	6.497

#1	.4107	.6125	.2882	.3459	.4071
#2	.5201	.7170	.4495	.3827	.4755
#3	.3498	.6149	.4750	.3682	.4492
#4	.4221	.6073	.3814	.4592	.4370
#5	.3729	.6787	.4251	.4699	.4774

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4211	.3646
Stddev	.0332	.0640
%RSD	7.889	17.54

#1	.3891	.2846
#2	.4235	.3418
#3	.4308	.3535
#4	.4703	.4592
#5	.3916	.3841

Sample Name: s98t003010 Run Time: 12/02/98 14:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	90.434
Stddev	1.240
%RSD	1.3708

#1	92.519
#2	89.722
#3	90.532
#4	90.016
#5	89.382

Method: ACTINIDE Sample Name: s98t003010_d

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 14:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa [231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3739	.3431	.4487	.7112	.3541
Stddev	.0454	.0261	.0458	.0660	.0277
%RSD	12.14	7.612	10.21	9.283	7.831
#1	.4428	.3147	.4282	.6039	.3184
#2	.3971	.3236	.3935	.7391	.3751
#3	.3464	.3391	.4448	.7693	.3792
#4	.3489	.3599	.4591	.6941	.3305
#5	.3343	.3784	.5180	.7495	.3674

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3995	.5998	.4370	.4340	.4165
Stddev	.0245	.0292	.0290	.0392	.0500
%RSD	6.123	4.864	6.646	9.034	12.00
#1	.4136	.6039	.4355	.3916	.4465
#2	.3898	.6398	.4744	.4118	.4780
#3	.4339	.5579	.4521	.4776	.3464
#4	.3709	.6023	.3966	.4150	.3966
#5	.3894	.5952	.4262	.4739	.4151

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4209	.4237
Stddev	.0411	.0223
%RSD	9.768	5.256
#1	.4794	.4062
#2	.4118	.4302
#3	.3828	.4521
#4	.3856	.3966
#5	.4445	.4335

Analysis Report

Sample Name: s98t003010_d Run Time: 12/02/98 14:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	90.929
Stddev	.319
%RSD	.35092

#1	91.078
#2	90.650
#3	91.420
#4	90.764
#5	90.732

Method: ACTINIDE Sample Name: s98t003011

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 14:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa [231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4154	.4076	.4795	.7119	.3959
Stddev	.0297	.0396	.0652	.0725	.0213
%RSD	7.139	9.725	13.60	10.18	5.378

#1	.4159	.3607	.4564	.6036	.4012
#2	.3726	.3801	.5105	.7043	.4173
#3	.4251	.4634	.5783	.7276	.3638
#4	.4548	.4215	.4326	.7173	.4104
#5	.4088	.4125	.4198	.8065	.3867

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4223	.6079	.4430	.4453	.4291
Stddev	.0180	.0422	.0232	.0595	.0384
%RSD	4.266	6.938	5.234	13.37	8.955

#1	.4343	.6220	.4711	.3496	.3975
#2	.4248	.6223	.4472	.4770	.3838
#3	.4442	.6548	.4481	.4404	.4710
#4	.4030	.5990	.4067	.4511	.4622
#5	.4051	.5414	.4419	.5082	.4309

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4415	.4510
Stddev	.0363	.0351
%RSD	8.222	7.790

#1	.4343	.4085
#2	.4583	.4695
#3	.4289	.4174
#4	.4918	.4770
#5	.3941	.4824

Sample Name: s98t003011 Run Time: 12/02/98 14:13

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	89.546
Stddev	1.468
%RSD	1.6388

#1	90.568
#2	89.455
#3	87.043
#4	90.153
#5	90.511

Method: ACTINIDE Sample Name: s98t003011_d

Operator: bjg

Comment: 40501 df

Run Time: 12/02/98 14:18 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pu[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4014	.3924	.4943	.7022	.4013
Stddev	.0274	.0428	.0329	.0658	.0194
%RSD	6.826	10.90	6.659	9.376	4.845
#1	.3793	.3574	.4960	.6710	.4303
#2	.3831	.3385	.5059	.8072	.4017
#3	.4034	.4108	.5403	.7217	.3997
#4	.4474	.4141	.4770	.6730	.3993
#5	.3938	.4412	.4521	.6381	.3756

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3987	.5856	.4428	.4124	.4208
Stddev	.0820	.0364	.0501	.0303	.0506
%RSD	20.58	6.208	11.31	7.352	12.02
#1	.3537	.5470	.4413	.4194	.3793
#2	.5319	.6435	.4761	.4533	.5022
#3	.3997	.5662	.4293	.3775	.4367
#4	.3143	.5916	.4992	.3883	.3993
#5	.3938	.5797	.3683	.4230	.3865

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4175	.4529
Stddev	.0208	.0766
%RSD	4.976	16.91
#1	.4267	.4012
#2	.3906	.4947
#3	.4034	.3442
#4	.4437	.5214
#5	.4230	.5032

Sample Name: s98t003011_d Run Time: 12/02/98 14:18

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	90.531
Stddev	.830
%RSD	.91636

#1	91.403
#2	89.613
#3	90.071
#4	90.148
#5	91.421

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 12/02/98 14:24 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa [231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3777	.4014	130.5	15.14	.6871
Stddev	.0607	.0186	5.5	.63	.0697
%RSD	16.07	4.636	4.191	4.170	10.15

#1	.4402	.3962	127.5	15.25	.7264
#2	.3504	.3754	127.4	15.27	.6424
#3	.4454	.4250	128.6	14.34	.7356
#4	.3354	.4122	128.7	14.80	.5860
#5	.3168	.3981	140.2	16.04	.7449

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	261.0	343.0	9.369	.4209	.4098
Stddev	8.1	10.8	.312	.0363	.0225
%RSD	3.113	3.163	3.325	8.735	5.487

#1	261.4	340.0	9.364	.3874	.4006
#2	252.3	335.1	8.919	.4714	.3754
#3	256.7	333.8	9.427	.3882	.4168
#4	260.4	345.4	9.340	.4122	.4324
#5	274.0	360.6	9.795	.4452	.4238

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1.072	.3862
Stddev	.066	.0368
%RSD	6.150	9.519

#1	1.008	.3786
#2	1.018	.3838
#3	1.132	.3800
#4	1.051	.3435
#5	1.152	.4452

Sample Name: ccv Run Time: 12/02/98 14:24

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	79.507
Stddev	2.753
%RSD	3.4626

#1	75.717
#2	79.905
#3	81.570
#4	82.479
#5	77.864

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 12/02/98 14:33 Type: Unk

Mode: IR

Corr.Fac: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3631	.3962	.4521	.4035	.3524
Stddev	.0361	.0361	.0488	.0229	.0487
%RSD	9.950	9.121	10.80	5.684	13.83
#1	.3146	.3854	.3854	.3972	.3540
#2	.4052	.3968	.4260	.3884	.2757
#3	.3489	.3965	.4520	.3886	.4005
#4	.3549	.3509	.5059	.3998	.3876
#5	.3920	.4514	.4911	.4435	.3445

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3842	.5870	.3974	.3915	.3824
Stddev	.0373	.0279	.0621	.0462	.0550
%RSD	9.700	4.751	15.63	11.80	14.37
#1	.3579	.6293	.3933	.3972	.4090
#2	.4469	.6015	.5054	.4635	.4177
#3	.3569	.5710	.3529	.3648	.4044
#4	.3713	.5712	.3713	.3917	.3957
#5	.3881	.5623	.3643	.3406	.2851

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3793	.3993
Stddev	.0552	.0584
%RSD	14.56	14.62
#1	.3146	.3894
#2	.3801	.4052
#3	.4163	.3807
#4	.4488	.3305
#5	.3366	.4911

Analysis Report

Sample Name: ccb Run Time: 12/02/98 14:33

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	82.900
Stddev	2.087
%RSD	2.5179

#1	84.754
#2	79.805
#3	84.065
#4	81.704
#5	84.173

HNF-1663 REV. 0
LABCORE Completed Worklist Report for Worklist# 27269

Analyst: bjg

Instrument: ICPMS1

Book#: 11B59

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) ACTINIDE

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSACT TH232	QC	1.0e-2	1.03e-02	103.000 %	Recovery
1	ICV	0	EMSACT NP237	QC	9.88e-3	1.14e-02	115.385 %	Recovery
1	ICV	0	EMSACT PU239	QC	1.00e-2	9.73e-03	97.300 %	Recovery
1	ICV	0	EMSACT 241	QC	1.026e-2	1.04e-02	101.365 %	Recovery
2	ICB	0	EMSACT1 TH229-A1	SOLID	1	<2.52e-6		ug/mL
2	ICB	0	EMSACT1 TH230-A1	SOLID	1	<1.38e-6		ug/mL
2	ICB	0	EMSACT1 TH232-A1	SOLID	1	<2.72e-6		ug/mL
3	BLNK-PREP	0	EMSACT1 PA231-A1	SOLID	1	<2.54e-5		ug/g
3	BLNK-PREP	0	EMSACT1 TH229-A1	SOLID	1	<2.52e-5		ug/g
3	BLNK-PREP	0	EMSACT1 TH230-A1	SOLID	1	<1.38e-5		ug/g
3	BLNK-PREP	0	EMSACT1 TH232-A1	SOLID	1	<2.72e-5		ug/g
3	BLNK-PREP	0	EMSACT1 NP237-A1	SOLID	1	<1.21e-5		ug/g
3	BLNK-PREP	0	EMSACT1 PU239-A1	SOLID	1	<3.79e-5		ug/g
3	BLNK-PREP	0	EMSACT1 PU240-A1	SOLID	1	<9.50e-6		ug/g
3	BLNK-PREP	0	EMSACT1 PU242-A1	SOLID	1	<7.15e-6		ug/g
3	BLNK-PREP	0	EMSACT1 241-A1	SOLID	1	<1.41e-5		ug/g
3	BLNK-PREP	0	EMSACT1 243-A1	SOLID	1	<2.06e-5		ug/g
3	BLNK-PREP	0	EMSACT1 244-A1	SOLID	1	<1.46e-5		ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 PA231-A1	SOLID	N/A	< 9.884e-02	9.88e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 TH229-A1	SOLID	N/A	< 9.832e-02	9.83e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 TH230-A1	SOLID	N/A	< 5.386e-02	5.39e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 TH232-A1	SOLID	N/A	< 2.774e+00	0.106	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 NP237-A1	SOLID	N/A	< 4.734e-02	4.73e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 PU239-A1	SOLID	N/A	< 1.476e-01	0.148	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 PU240-A1	SOLID	N/A	< 3.703e-02	3.70e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 PU242-A1	SOLID	N/A	< 2.786e-02	2.79e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 241-A1	SOLID	N/A	< 5.492e-02	5.49e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 243-A1	SOLID	N/A	< 8.046e-02	8.05e-002	ug/g
4	SAMPLE	S98T002990	0 A EMSACT1 244-A1	SOLID	N/A	< 5.701e-02	5.70e-002	ug/g
5	DUP	S98T002990	0 A EMSACT1 PA231-A1	SOLID	<9.88e-2	<1.01e-1		RPD
5	DUP	S98T002990	0 A EMSACT1 TH229-A1	SOLID	<9.83e-2	<1.01e-1		RPD
5	DUP	S98T002990	0 A EMSACT1 TH230-A1	SOLID	<5.39e-2	<5.52e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 TH232-A1	SOLID	2.77e+00	2.08e+00	28.454	RPD
5	DUP	S98T002990	0 A EMSACT1 NP237-A1	SOLID	<4.73e-2	<4.86e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 PU239-A1	SOLID	<1.48e-1	<1.51e-1		RPD
5	DUP	S98T002990	0 A EMSACT1 PU240-A1	SOLID	<3.70e-2	<3.80e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 PU242-A1	SOLID	<2.79e-2	<2.86e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 241-A1	SOLID	<5.49e-2	<5.63e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 243-A1	SOLID	<8.05e-2	<8.25e-2		RPD
5	DUP	S98T002990	0 A EMSACT1 244-A1	SOLID	<5.70e-2	<5.85e-2		RPD
6	SPK-POST	S98T002990	0 A EMSACT1 TH232-A1	SOLID	0.020	1.97e-02	98.500 %	Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27269

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SPK-POST	S98T002990	0 A	OMSACTA1 NP237-A1	SOLID	0.01976	2.24e-02	113.360	% Recovery
6 SPK-POST	S98T002990	0 A	OMSACTA1 PU239-A1	SOLID	0.02008	1.96e-02	97.610	% Recovery
6 SPK-POST	S98T002990	0 A	OMSACTA1 241-A1	SOLID	0.02052	2.10e-02	102.339	% Recovery
7 SAMPLE	S98T002991	0 A	OMSACTA1 PA231-A1	SOLID	N/A	< 1.010e-01	0.101	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 TH229-A1	SOLID	N/A	< 1.005e-01	0.101	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 TH230-A1	SOLID	N/A	< 5.505e-02	5.51e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 TH232-A1	SOLID	N/A	< 3.072e+00	0.108	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 NP237-A1	SOLID	N/A	< 4.839e-02	4.84e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 PU239-A1	SOLID	N/A	< 2.085e-01	0.151	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 PU240-A1	SOLID	N/A	< 3.785e-02	3.79e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 PU242-A1	SOLID	N/A	< 2.848e-02	2.85e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 241-A1	SOLID	N/A	< 5.614e-02	5.61e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 243-A1	SOLID	N/A	< 8.225e-02	8.23e-002	ug/g
7 SAMPLE	S98T002991	0 A	OMSACTA1 244-A1	SOLID	N/A	< 5.828e-02	5.83e-002	ug/g
8 DUP	S98T002991	0 A	OMSACTA1 PA231-A1	SOLID	<1.01e-1	<1.03e-1		RPD
8 DUP	S98T002991	0 A	OMSACTA1 TH229-A1	SOLID	<1.01e-1	<1.03e-1		RPD
8 DUP	S98T002991	0 A	OMSACTA1 TH230-A1	SOLID	<5.51e-2	<5.64e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 TH232-A1	SOLID	3.07e+00	2.70e+00	12.825	RPD
8 DUP	S98T002991	0 A	OMSACTA1 NP237-A1	SOLID	<4.84e-2	<4.96e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 PU239-A1	SOLID	2.08e-01	<1.55e-1		RPD
8 DUP	S98T002991	0 A	OMSACTA1 PU240-A1	SOLID	<3.79e-2	<3.88e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 PU242-A1	SOLID	<2.85e-2	<2.92e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 241-A1	SOLID	<5.61e-2	<5.75e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 243-A1	SOLID	<8.23e-2	<8.43e-2		RPD
8 DUP	S98T002991	0 A	OMSACTA1 244-A1	SOLID	<5.83e-2	<5.97e-2		RPD
9 SPK-POST	S98T002991	0 A	OMSACTA1 TH232-A1	SOLID	0.020	1.91e-02	95.500	% Recovery
9 SPK-POST	S98T002991	0 A	OMSACTA1 NP237-A1	SOLID	0.01976	2.18e-02	110.324	% Recovery
9 SPK-POST	S98T002991	0 A	OMSACTA1 PU239-A1	SOLID	0.02008	1.90e-02	94.622	% Recovery
9 SPK-POST	S98T002991	0 A	OMSACTA1 241-A1	SOLID	0.02052	2.04e-02	99.415	% Recovery
10 CCV	0	0	OMSACT TH232	QC	1.00e-2	1.04e-02	104.000	% Recovery
10 CCV	0	0	OMSACT NP237	QC	9.88e-3	1.15e-02	116.397	% Recovery
10 CCV	0	0	OMSACT PU239	QC	1.004e-2	1.00e-02	99.602	% Recovery
10 CCV	0	0	OMSACT 241	QC	1.026e-2	1.08e-02	105.263	% Recovery
11 CCB	0	0	OMSACTA1 PA231-A1	LIQUID	1	<2.54e-6		ug/mL
11 CCB	0	0	OMSACTA1 TH229-A1	LIQUID	1	<2.52e-6		ug/mL
11 CCB	0	0	OMSACTA1 TH230-A1	LIQUID	1	<1.38e-6		ug/mL
11 CCB	0	0	OMSACTA1 TH232-A1	LIQUID	1	<2.72e-6		ug/mL
11 CCB	0	0	OMSACTA1 PU239-A1	LIQUID	1	<3.79e-6		ug/mL
11 CCB	0	0	OMSACTA1 PU240-A1	LIQUID	1	<9.50e-7		ug/mL
11 CCB	0	0	OMSACTA1 PU242-A1	LIQUID	1	<7.15e-7		ug/mL
11 CCB	0	0	OMSACTA1 241-A1	LIQUID	1	<1.41e-6		ug/mL
11 CCB	0	0	OMSACTA1 243-A1	LIQUID	1	<2.06e-6		ug/mL
11 CCB	0	0	OMSACTA1 244-A1	LIQUID	1	<1.46e-6		ug/mL
12 SAMPLE	S98T002992	0 A	OMSACTA1 PA231-A1	SOLID	N/A	< 9.958e-02	9.96e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 TH229-A1	SOLID	N/A	< 9.905e-02	9.90e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 TH230-A1	SOLID	N/A	< 5.426e-02	5.42e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 TH232-A1	SOLID	N/A	< 2.927e+00	0.107	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 NP237-A1	SOLID	N/A	< 4.769e-02	4.77e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 PU239-A1	SOLID	N/A	< 1.487e-01	0.149	ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27269

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
12 SAMPLE	S98T002992	0 A	OMSACTA1 PU240-A1	SOLID	N/A <	3.730e-02	3.73e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 PU242-A1	SOLID	N/A <	2.807e-02	2.81e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 241-A1	SOLID	N/A <	5.533e-02	5.53e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 243-A1	SOLID	N/A <	8.106e-02	8.11e-002	ug/g
12 SAMPLE	S98T002992	0 A	OMSACTA1 244-A1	SOLID	N/A <	5.744e-02	5.74e-002	ug/g
13 DUP	S98T002992	0 A	OMSACTA1 PA231-A1	SOLID	<9.96e-2	<1.01e-1		RPD
13 DUP	S98T002992	0 A	OMSACTA1 TH229-A1	SOLID	<9.90e-2	<1.01e-1		RPD
13 DUP	S98T002992	0 A	OMSACTA1 TH230-A1	SOLID	<5.43e-2	<5.51e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 TH232-A1	SOLID	2.93e+00	2.28e+00	24.952	RPD
13 DUP	S98T002992	0 A	OMSACTA1 NP237-A1	SOLID	<4.77e-2	<4.84e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 PU239-A1	SOLID	<1.49e-1	<1.51e-1		RPD
13 DUP	S98T002992	0 A	OMSACTA1 PU240-A1	SOLID	<3.73e-2	<3.78e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 PU242-A1	SOLID	<2.81e-2	<2.85e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 241-A1	SOLID	<5.53e-2	<5.61e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 243-A1	SOLID	<8.11e-2	<8.22e-2		RPD
13 DUP	S98T002992	0 A	OMSACTA1 244-A1	SOLID	<5.74e-2	<5.83e-2		RPD
14 SAMPLE	S98T002993	0 A	OMSACTA1 PA231-A1	SOLID	N/A <	1.016e-01	0.102	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 TH229-A1	SOLID	N/A <	1.011e-01	0.101	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 TH230-A1	SOLID	N/A <	5.538e-02	5.54e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 TH232-A1	SOLID	N/A <	1.325e+00	0.109	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 NP237-A1	SOLID	N/A <	4.868e-02	4.87e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 PU239-A1	SOLID	N/A <	1.518e-01	0.152	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 PU240-A1	SOLID	N/A <	3.808e-02	3.81e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 PU242-A1	SOLID	N/A <	2.865e-02	2.86e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 241-A1	SOLID	N/A <	5.647e-02	5.65e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 243-A1	SOLID	N/A <	8.274e-02	8.28e-002	ug/g
14 SAMPLE	S98T002993	0 A	OMSACTA1 244-A1	SOLID	N/A <	5.863e-02	5.86e-002	ug/g
15 DUP	S98T002993	0 A	OMSACTA1 PA231-A1	SOLID	<1.02e-1	<1.05e-1		RPD
15 DUP	S98T002993	0 A	OMSACTA1 TH229-A1	SOLID	<1.01e-1	<1.04e-1		RPD
15 DUP	S98T002993	0 A	OMSACTA1 TH230-A1	SOLID	<5.54e-2	<5.70e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 TH232-A1	SOLID	1.32e+00	1.43e+00	8.000	RPD
15 DUP	S98T002993	0 A	OMSACTA1 NP237-A1	SOLID	<4.87e-2	<5.01e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 PU239-A1	SOLID	<1.52e-1	<1.56e-1		RPD
15 DUP	S98T002993	0 A	OMSACTA1 PU240-A1	SOLID	<3.81e-2	<3.92e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 PU242-A1	SOLID	<2.86e-2	<2.95e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 241-A1	SOLID	<5.65e-2	<5.81e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 243-A1	SOLID	<8.27e-2	<8.52e-2		RPD
15 DUP	S98T002993	0 A	OMSACTA1 244-A1	SOLID	<5.86e-2	<6.03e-2		RPD
16 SPK-POST	S98T002993	0 A	OMSACTA1 TH232-A1	SOLID	0.020	2.00e-02	100.000 % Recovery	
16 SPK-POST	S98T002993	0 A	OMSACTA1 NP237-A1	SOLID	0.019760	2.24e-02	113.360 % Recovery	
16 SPK-POST	S98T002993	0 A	OMSACTA1 PU239-A1	SOLID	0.020080	1.95e-02	97.112 % Recovery	
16 SPK-POST	S98T002993	0 A	OMSACTA1 241-A1	SOLID	0.020520	2.11e-02	102.827 % Recovery	
17 CCV	0		OMSAC-QC TH232	QC	1.00e-2	1.05e-02	105.000 % Recovery	
17 CCV	0		OMSAC-QC NP237	QC	9.88e-3	1.15e-02	116.397 % Recovery	
17 CCV	0		OMSAC-QC PU239	QC	1.05e-2	1.00e-02	95.238 % Recovery	
17 CCV	0		OMSAC-QC 241	QC	1.004e-2	1.09e-02	108.566 % Recovery	
18 CCB	0		OMSACTA1 PA231-A1	LIQUID	1	<2.54e-6		ug/mL
18 CCB	0		OMSACTA1 TH229-A1	LIQUID	1	<2.52e-6		ug/mL
18 CCB	0		OMSACTA1 TH230-A1	LIQUID	1	<1.38e-6		ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27269

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
18 CCB	0	@MSACTA1 TH232-A1	LIQUID	1	<2.72e-6		ug/mL
18 CCB	0	@MSACTA1 PU239-A1	LIQUID	1	<3.79e-6		ug/mL
18 CCB	0	@MSACTA1 PU240-A1	LIQUID	1	<9.50e-7		ug/mL
18 CCB	0	@MSACTA1 PU242-A1	LIQUID	1	<7.15e-7		ug/mL
18 CCB	0	@MSACTA1 241-A1	LIQUID	1	<1.41e-6		ug/mL
18 CCB	0	@MSACTA1 243-A1	LIQUID	1	<2.06e-6		ug/mL
18 CCB	0	@MSACTA1 244-A1	LIQUID	1	<1.46e-6		ug/mL

Final page for worklist# 27269

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul M. Pang 03/03/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27269

Analyst: BJG Instrument: ICPMS1 _____ Book# 11859Method: LA-506-101 Rev/Mod AO

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) ACTINIDE

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICB			@MSACTA1	SOLID		
3 BLNK-PREP			@MSACTA1	SOLID		
4 SAMPLE	S98T002990 0 A		@MSACTA1	SOLID	98000509	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
5 DUP	S98T002990 0 A		@MSACTA1	SOLID		
6 SPK- ^{post}	S98T002990 0 A		@MSACTA1	SOLID		
7 SAMPLE ^{9/3/99}	S98T002991 0 A		@MSACTA1	SOLID	98000509	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
8 DUP	S98T002991 0 A		@MSACTA1	SOLID		
9 SPK- ^{post}	S98T002991 0 A		@MSACTA1	SOLID		
10 CCV ^{9/3/99}			@MSACT	QC		
11 CCB			@MSACTA1	SOLID		
12 SAMPLE	S98T002992 0 A		@MSACTA1	SOLID	98000509	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
13 DUP	S98T002992 0 A		@MSACTA1	SOLID		
14 SAMPLE	S98T002993 0 A		@MSACTA1	SOLID	98000509	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
15 DUP	S98T002993 0 A		@MSACTA1	SOLID		

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

03/01/99 10:25

Page: 2

ws2

LABCORE Data Entry Template for Worklist# 27269

S Type	Sample#	R	A	Test	Matrix	Group#	Project
16 SPK-post	S98T002993	0	A	@MSACTA1	SOLID		
17 CCV	982/3/99			@MSACT	QC		
18 CCB				@MSACTA1	SOLID		

Final page for worklist # 27269

B. K. S. 3-0-99
 Signature Date

Signature Date

PREPBLK 1st - 2nd
 S98T002990 .05ml - 10ml
 2990-D
 2990-A
 4/3-199 ~~2990-S~~
 S98T002991
 2991-D
 2991-A
 4/3-199 ~~2991-S~~
 S98T002992
 2992-D
 S98T002993
 2993-D
 2993-A
 4/3-199 ~~2993-S~~

DF
 10
 201

Samples previously scanned.
 POST SPIKES - .05ml of MSACT. Spk 8859

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2155

Method: MS_TUNE Sample Name: ms-tune Operator: bjg
Comment:
Run Time: 03/01/99 19:43 Type: Unk Mode: INT Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	232200.
Stddev	1913.
%RSD	.8240

#1	232800.
#2	234600.
#3	232800.
#4	233000.
#5	229300.
#6	231500.
#7	234100.
#8	229100.
#9	233800.
#10	231000.

bjg

3-1-99

ACTAIDE

BY-105

5587002980-93

FILE# AC03014.TFT

W.L. # 27269

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2156 TO 2198.

Method: ACTINIDE Sample Name: cal blank

Operator: bjg

Comment: background

Run Time: 03/01/99 19:50 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2248	.2546	.2947	.2962	.2707
Stddev	.0318	.0177	.0359	.0315	.0330
%RSD	14.15	6.953	12.17	10.63	12.17

#1	.2702	.2754	.2492	.2413	.2728
#2	.2397	.2502	.2950	.3108	.2686
#3	.2070	.2547	.2972	.3211	.2176
#4	.1870	.2645	.2832	.3046	.3046
#5	.2200	.2281	.3488	.3032	.2898

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2330	.2737	.2261	.2388	.2462
Stddev	.0121	.0388	.0141	.0109	.0229
%RSD	5.210	14.16	6.249	4.553	9.311

#1	.2256	.2675	.2072	.2203	.2570
#2	.2318	.3029	.2291	.2449	.2634
#3	.2176	.2415	.2362	.2468	.2096
#4	.2431	.3233	.2164	.2378	.2378
#5	.2469	.2335	.2415	.2442	.2630

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2457	.2451
Stddev	.0212	.0319
%RSD	8.621	13.02

#1	.2466	.2728
#2	.2370	.2265
#3	.2149	.2149
#4	.2619	.2859
#5	.2683	.2254

2157

HNF-1663 REV. 0

Sample Name: cal blank Run Time: 03/01/99 19:50

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	125.65
Stddev	1.20
%RSD	.95386

#1	127.09
#2	126.56
#3	125.61
#4	124.75
#5	124.22

2158

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.010 ppm cal

Run Time: 03/01/99 19:56 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1668	.2508	925.6	1011.	1.318
Stddev	.0354	.0393	9.0	6.	.050
%RSD	21.19	15.66	.9727	.5565	3.817
#1	.1798	.2781	911.2	1002.	1.314
#2	.1373	.1862	929.6	1015.	1.310
#3	.1331	.2479	926.5	1014.	1.341
#4	.1641	.2563	925.1	1011.	1.243
#5	.2199	.2857	935.6	1015.	1.381

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.574	.3619	.1595	.1675	.1616
Stddev	.019	.0294	.0026	.0321	.0121
%RSD	1.184	8.131	1.642	19.14	7.510
#1	1.589	.3140	.1558	.1990	.1726
#2	1.594	.3537	.1606	.1885	.1745
#3	1.549	.3788	.1630	.1813	.1584
#4	1.576	.3867	.1596	.1461	.1574
#5	1.562	.3763	.1587	.1224	.1451

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1616	.1493
Stddev	.0298	.0222
%RSD	18.41	14.87
#1	.1894	.1750
#2	.1559	.1420
#3	.1217	.1699
#4	.1484	.1236
#5	.1927	.1360

Sample Name: cal std Run Time: 03/01/99 19:56

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	144.56
Stddev	3.62
%RSD	2.5051
#1	139.05
#2	143.24
#3	145.21
#4	148.28
#5	147.03

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 03/01/99 20:01 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2506	1.060	8952.	9602.	11.63
Stddev	.0239	.060	121.	71.	.10
%RSD	9.519	5.682	1.357	.7411	.8546
#1	.2416	1.017	8808.	9676.	11.69
#2	.2306	1.030	8861.	9589.	11.63
#3	.2306	1.103	8970.	9553.	11.57
#4	.2651	1.008	9115.	9674.	11.75
#5	.2850	1.145	9007.	9518.	11.50

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	14.13	2.030	.1532	.1522	.1622
Stddev	.20	.133	.0126	.0125	.0153
%RSD	1.418	6.561	8.228	8.217	9.437
#1	14.08	2.117	.1595	.1573	.1550
#2	14.15	2.087	.1621	.1461	.1415
#3	13.85	1.802	.1644	.1393	.1644
#4	14.41	2.025	.1440	.1714	.1829
#5	14.15	2.120	.1357	.1470	.1674

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1645	.1755
Stddev	.0266	.0226
%RSD	16.18	12.87
#1	.1595	.1573
#2	.1895	.1803
#3	.1256	.1599
#4	.1897	.2126
#5	.1583	.1674

Sample Name: cal std Run Time: 03/01/99 20:01

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	146.28
Stddev	.62
%RSD	.42418

#1	146.24
#2	146.02
#3	145.97
#4	145.83
#5	147.36

2162

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 03/01/99 20:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2241	.2956	956.6	15.42	1.459
Stddev	.0095	.0136	19.2	.24	.076
%RSD	4.253	4.588	2.011	1.574	5.235
#1	.2179	.2865	937.3	15.41	1.423
#2	.2109	.3101	935.6	15.25	1.427
#3	.2311	.3090	970.9	15.72	1.595
#4	.2267	.2796	977.5	15.58	1.428
#5	.2338	.2929	961.6	15.12	1.420

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1135.	997.5	26.63	.3403	.2088
Stddev	15.	19.4	.54	.0143	.0204
%RSD	1.323	1.943	2.023	4.197	9.786
#1	1116.	979.3	25.87	.3306	.2057
#2	1137.	1018.	26.38	.3225	.2134
#3	1146.	1015.	27.20	.3517	.2336
#4	1153.	976.5	27.06	.3400	.2141
#5	1124.	998.4	26.66	.3569	.1772

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1094.	.2349
Stddev	26.	.0200
%RSD	2.349	8.497
#1	1066.	.2302
#2	1073.	.2679
#3	1120.	.2135
#4	1121.	.2317
#5	1092.	.2314

Sample Name: icv Run Time: 03/01/99 20:13

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	134.19
Stddev	1.65
%RSD	1.2303
#1	136.13
#2	134.36
#3	132.71
#4	132.34
#5	135.42

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 20:29 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2051	.2046	.3555	.2601	.2090
Stddev	.0249	.0200	.0305	.0231	.0084
%RSD	12.14	9.764	8.580	8.873	4.021

#1	.2385	.1770	.3983	.2705	.1967
#2	.2043	.1920	.3274	.2462	.2043
#3	.1685	.2106	.3295	.2651	.2156
#4	.2097	.2270	.3478	.2294	.2122
#5	.2043	.2165	.3746	.2895	.2165

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2174	.2524	.2130	.2065	.2062
Stddev	.0116	.0242	.0190	.0196	.0218
%RSD	5.322	9.600	8.920	9.476	10.57

#1	.2237	.2729	.2410	.1819	.1918
#2	.2043	.2585	.2018	.1945	.2092
#3	.2056	.2230	.2007	.2131	.2205
#4	.2294	.2763	.2245	.2097	.2319
#5	.2238	.2311	.1970	.2335	.1776

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2007	.2252
Stddev	.0267	.0283
%RSD	13.30	12.54

#1	.2311	.2459
#2	.2265	.2585
#3	.1933	.1982
#4	.1801	.1949
#5	.1727	.2287

2165

Sample Name: ich Run Time: 03/01/99 20:29

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	135.53
Stddev	.92
%RSD	.68102

#1	135.57
#2	135.41
#3	134.53
#4	135.12
#5	137.02

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/01/99 20:39 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1955	.1831	.4333	.5108	.2121
Stddev	.0170	.0269	.0265	.0259	.0226
%RSD	8.715	14.69	6.117	5.079	10.65

#1	.1840	.1613	.4528	.5207	.2207
#2	.1713	.1713	.4201	.4844	.2141
#3	.2071	.1964	.4331	.5030	.2233
#4	.2043	.1624	.4637	.4951	.1729
#5	.2106	.2240	.3968	.5507	.2294

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2106	.2810	.2174	.1984	.2165
Stddev	.0216	.0299	.0226	.0310	.0202
%RSD	10.27	10.65	10.41	15.64	9.345

#1	.1981	.3340	.2066	.2490	.2406
#2	.1873	.2703	.2248	.1820	.2114
#3	.2044	.2609	.1910	.1910	.2313
#4	.2201	.2725	.2515	.1677	.1886
#5	.2429	.2672	.2133	.2025	.2106

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2019	.2070
Stddev	.0171	.0150
%RSD	8.464	7.269

#1	.1868	.2123
#2	.2301	.1820
#3	.1964	.2044
#4	.2043	.2174
#5	.1917	.2186

2167

Sample Name: prepblk Run Time: 03/01/99 20:39

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	123.40
Stddev	3.46
%RSD	2.8042

#1	117.78
#2	124.56
#3	123.91
#4	127.24
#5	123.49

2168

Method: ACTINIDE Sample Name: s98t002990

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 20:44 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1635	.1736	6.878	420.3	.1909
Stddev	.0270	.0115	.146	6.4	.0258
%RSD	16.49	6.641	2.129	1.530	13.53

#1	.1932	.1765	6.798	422.0	.1765
#2	.1852	.1757	6.797	417.6	.1780
#3	.1415	.1559	6.873	425.6	.1631
#4	.1665	.1879	6.789	410.5	.2188
#5	.1309	.1721	7.132	425.9	.2182

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8192	.3588	.1865	.1813	.1894
Stddev	.0280	.0336	.0106	.0025	.0338
%RSD	3.422	9.373	5.663	1.361	17.85

#1	.8158	.3602	.1884	.1837	.2314
#2	.7905	.3181	.2042	.1804	.1662
#3	.8180	.3383	.1799	.1775	.1487
#4	.8063	.4067	.1784	.1832	.2141
#5	.8655	.3709	.1818	.1818	.1867

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2023	.1795
Stddev	.0182	.0285
%RSD	8.978	15.88

#1	.2314	.1837
#2	.1970	.1733
#3	.1847	.1439
#4	.2069	.1736
#5	.1915	.2230

Sample Name: s98t002990 Run Time: 03/01/99 20:44

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	139.35
Stddev	1.17
%RSD	.84268

#1	139.74
#2	140.41
#3	138.95
#4	140.14
#5	137.49

Method: ACTINIDE

Sample Name: s98t002990_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 20:49 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1843	.1654	5.117	301.8	.1794
Stddev	.0151	.0250	.116	2.5	.0202
%RSD	8.199	15.10	2.260	.8158	11.28
#1	.2086	.1988	5.263	304.7	.2111
#2	.1683	.1854	5.095	302.3	.1561
#3	.1865	.1477	5.088	303.0	.1816
#4	.1806	.1517	4.955	298.3	.1782
#5	.1775	.1435	5.185	300.6	.1702

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.6199	.3116	.1838	.1755	.1667
Stddev	.0115	.0323	.0302	.0229	.0201
%RSD	1.857	10.37	16.42	13.02	12.06
#1	.6357	.2798	.1988	.1767	.1375
#2	.6073	.3488	.1732	.2122	.1780
#3	.6127	.2809	.1356	.1598	.1574
#4	.6163	.3082	.2094	.1757	.1709
#5	.6274	.3405	.2018	.1532	.1897

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1746	.1794
Stddev	.0226	.0195
%RSD	12.93	10.90
#1	.2135	.1939
#2	.1683	.1512
#3	.1550	.1695
#4	.1709	.1998
#5	.1654	.1824

Sample Name: s98t002990_d Run Time: 03/01/99 20:49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	137.13
Stddev	1.00
%RSD	.73022

#1	135.80
#2	136.67
#3	137.64
#4	138.46
#5	137.07

Method: ACTINIDE Sample Name: s98t002990_a

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 20:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2135	.3504	1829.	430.8	2.431
Stddev	.0301	.0311	13.	2.5	.063
%RSD	14.09	8.866	.7026	.5804	2.573
#1	.2134	.2992	1839.	430.3	2.450
#2	.1740	.3803	1819.	430.2	2.421
#3	.2496	.3669	1841.	434.1	2.508
#4	.1960	.3473	1836.	432.2	2.441
#5	.2348	.3583	1813.	427.4	2.335

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2230.	2009.	54.37	.3959	.2041
Stddev	8.	19.	.41	.0543	.0135
%RSD	.3754	.9654	.7612	13.72	6.617
#1	2218.	1977.	53.97	.4243	.2183
#2	2229.	2012.	54.10	.3753	.1939
#3	2239.	2029.	54.77	.4742	.2022
#4	2236.	2014.	54.85	.3722	.1886
#5	2225.	2013.	54.13	.3336	.2175

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2214.	.2259
Stddev	8.	.0183
%RSD	.3486	8.099
#1	2207.	.2085
#2	2220.	.2063
#3	2224.	.2271
#4	2210.	.2431
#5	2208.	.2447

Sample Name: s98t002990_a Run Time: 03/01/99 20:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	134.56
Stddev	.90
%RSD	.66520

#1	135.92
#2	134.10
#3	133.56
#4	134.35
#5	134.88

HNF-1663 REV.0

Method: ACTINIDE Sample Name: s98t002991

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1709	.1783	7.429	566.5	.1888
Stddev	.0219	.0215	.214	6.9	.0226
%RSD	12.80	12.06	2.887	1.217	11.97

#1	.1349	.1569	7.660	572.0	.1716
#2	.1719	.1599	7.274	557.1	.1862
#3	.1839	.1889	7.547	569.2	.1668
#4	.1920	.2090	7.139	572.7	.2236
#5	.1719	.1767	7.524	561.4	.1958

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.9909	1.142	.2081	.1953	.1759
Stddev	.0521	.096	.0243	.0321	.0171
%RSD	5.258	8.384	11.68	16.42	9.702

#1	1.049	1.282	.1839	.1888	.1961
#2	.9382	1.148	.2387	.2053	.1838
#3	1.001	1.172	.1889	.2428	.1790
#4	1.030	1.052	.2284	.1555	.1701
#5	.9360	1.053	.2006	.1839	.1504

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2038	.1898
Stddev	.0357	.0206
%RSD	17.54	10.84

#1	.2305	.2035
#2	.2531	.2172
#3	.1766	.1815
#4	.1750	.1822
#5	.1839	.1648

2175

Sample Name: s98t002991 Run Time: 03/01/99 21:02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	137.65
Stddev	1.86
%RSD	1.3518

#1	135.95
#2	139.63
#3	135.91
#4	137.18
#5	139.60

Method: ACTINIDE

Sample Name: s98t002991_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1859	.1660	6.416	536.1	.1830
Stddev	.0151	.0175	.131	3.5	.0143
%RSD	8.096	10.57	2.048	.6580	7.801
#1	.2048	.1524	6.411	539.6	.1596
#2	.1784	.1617	6.607	537.1	.1903
#3	.1792	.1792	6.267	530.3	.1792
#4	.1984	.1888	6.468	537.8	.1936
#5	.1688	.1477	6.329	535.5	.1923

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.9976	.8777	.1925	.1783	.1830
Stddev	.0591	.0556	.0286	.0152	.0226
%RSD	5.921	6.333	14.87	8.527	12.36
#1	.9407	.9288	.1548	.1905	.1524
#2	1.075	.9466	.1784	.1665	.1712
#3	1.021	.8488	.2028	.1603	.1981
#4	1.016	.8342	.2318	.1960	.2103
#5	.9356	.8301	.1946	.1782	.1829

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1844	.1641
Stddev	.0064	.0160
%RSD	3.489	9.780
#1	.1881	.1596
#2	.1903	.1903
#3	.1886	.1533
#4	.1793	.1673
#5	.1759	.1501

2177

Sample Name: s98t002991_d Run Time: 03/01/99 21:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	140.62
Stddev	1.11
%RSD	.78776

#1	139.96
#2	140.15
#3	141.37
#4	139.46
#5	142.15

Method: ACTINIDE

Sample Name: s98t002991_a

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:12 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1877	.3270	1770.	575.2	2.334
Stddev	.0132	.0161	16.	2.3	.093
%RSD	7.046	4.924	.9048	.4010	4.001
#1	.1744	.3346	1753.	579.0	2.268
#2	.1824	.3121	1764.	575.7	2.489
#3	.1970	.3508	1794.	573.9	2.285
#4	.1787	.3241	1777.	573.2	2.274
#5	.2058	.3135	1762.	574.2	2.353

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2175.	1949.	52.76	.3749	.1771
Stddev	7.	13.	.88	.0546	.0121
%RSD	.3039	.6595	1.673	14.56	6.813
#1	2173.	1927.	51.36	.3393	.1840
#2	2169.	1948.	52.45	.3817	.1752
#3	2183.	1956.	53.22	.3532	.1634
#4	2168.	1958.	53.19	.3336	.1692
#5	2180.	1955.	53.58	.4667	.1939

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2146.	.1915
Stddev	5.	.0247
%RSD	.2412	12.88
#1	2149.	.1673
#2	2141.	.1920
#3	2153.	.2235
#4	2141.	.2073
#5	2148.	.1675

Sample Name: s98t002991_a   Run Time: 03/01/99 21:12

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	139.24
Stddev	.46
%RSD	.33345
#1	139.49
#2	138.87
#3	138.73
#4	139.87
#5	139.27

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/01/99 21:23 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2046	.2714	964.9	15.74	1.327
Stddev	.0151	.0150	12.0	.21	.035
%RSD	7.382	5.536	1.246	1.358	2.636
#1	.1981	.2506	962.4	15.51	1.277
#2	.1847	.2807	982.4	16.03	1.334
#3	.2092	.2886	965.7	15.56	1.311
#4	.2053	.2626	948.7	15.76	1.342
#5	.2259	.2744	965.1	15.85	1.370

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1141.	1025.	27.81	.2776	.1845
Stddev	8.	7.	.33	.0295	.0370
%RSD	.7219	.6653	1.170	10.64	20.02
#1	1131.	1029.	27.74	.2673	.2052
#2	1132.	1016.	27.51	.2327	.1631
#3	1147.	1022.	27.60	.3078	.1996
#4	1146.	1024.	27.89	.2984	.1313
#5	1147.	1033.	28.34	.2817	.2234

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1139.	.2215
Stddev	7.	.0236
%RSD	.6357	10.64
#1	1141.	.2052
#2	1133.	.2447
#3	1133.	.2333
#4	1136.	.1886
#5	1151.	.2356

2181

Sample Name: ccv Run Time: 03/01/99 21:23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	138.83
Stddev	.99
%RSD	.71661
#1	139.69
#2	138.96
#3	138.60
#4	139.63
#5	137.25

2182

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 21:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1922	.1875	.4022	.2517	.1965
Stddev	.0194	.0129	.0298	.0364	.0313
%RSD	10.12	6.865	7.419	14.45	15.91

#1	.1799	.1993	.3987	.2285	.1532
#2	.2099	.1694	.4509	.2767	.2004
#3	.1962	.1986	.4044	.2872	.1843
#4	.1653	.1797	.3737	.2659	.2060
#5	.2097	.1904	.3832	.2000	.2386

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2018	.5402	.2071	.1899	.1903
Stddev	.0114	.0271	.0123	.0136	.0169
%RSD	5.634	5.023	5.953	7.186	8.897

#1	.1993	.5689	.2042	.2091	.1848
#2	.2219	.5606	.1909	.1789	.1932
#3	.1962	.5480	.2154	.1915	.1651
#4	.1940	.5078	.2228	.1749	.2108
#5	.1976	.5158	.2025	.1952	.1976

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2149	.1985
Stddev	.0221	.0189
%RSD	10.30	9.539

#1	.2407	.1993
#2	.2147	.1742
#3	.2178	.1891
#4	.1797	.2252
#5	.2217	.2049

Sample Name: ccb Run Time: 03/01/99 21:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	138.72
Stddev	1.03
%RSD	.74557

#1	137.11
#2	139.73
#3	139.29
#4	139.16
#5	138.30

Analysis Report

Method: ACTINIDE Sample Name: s98t002992

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:35 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1770	.1834	7.190	445.9	.1750
Stddev	.0267	.0110	.096	4.4	.0261
%RSD	15.09	6.001	1.336	.9916	14.90

#1	.1750	.1870	7.321	449.2	.1654
#2	.2077	.1695	7.106	442.3	.1791
#3	.1765	.1765	7.184	449.5	.2171
#4	.1903	.1856	7.246	448.4	.1475
#5	.1354	.1985	7.093	440.0	.1658

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8384	.6715	.1967	.1840	.1835
Stddev	.0343	.0521	.0153	.0169	.0199
%RSD	4.094	7.762	7.797	9.167	10.87

#1	.8103	.7528	.1798	.1942	.2134
#2	.8572	.6685	.1838	.1982	.1576
#3	.7967	.6750	.2051	.1598	.1789
#4	.8802	.6518	.1974	.1951	.1856
#5	.8476	.6094	.2171	.1728	.1821

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1935	.1788
Stddev	.0156	.0228
%RSD	8.042	12.77

#1	.2182	.1654
#2	.1862	.1862
#3	.1861	.2123
#4	.1784	.1784
#5	.1985	.1518

Sample Name: s98t002992 Run Time: 03/01/99 21:35

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	140.26
Stddev	1.45
%RSD	1.0360
#1	139.04
#2	139.61
#3	139.75
#4	140.12
#5	142.76

2186

Method: ACTINIDE

Sample Name: s98t002992_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:40 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1722	.1469	5.586	377.5	.1717
Stddev	.0227	.0214	.172	2.8	.0218
%RSD	13.18	14.57	3.082	.7427	12.73

#1	.1728	.1586	5.562	381.2	.1609
#2	.2048	.1553	5.559	376.6	.2024
#3	.1784	.1367	5.454	374.5	.1668
#4	.1621	.1690	5.474	375.7	.1829
#5	.1429	.1148	5.882	379.6	.1452

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.7268	.5926	.1865	.1819	.1989
Stddev	.0425	.0659	.0244	.0191	.0214
%RSD	5.842	11.12	13.10	10.51	10.76

#1	.7077	.6816	.1775	.1704	.2154
#2	.6825	.6143	.1577	.1695	.2236
#3	.6998	.6001	.1900	.1877	.1691
#4	.7664	.5024	.2246	.1690	.1945
#5	.7776	.5645	.1827	.2131	.1921

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1731	.1660
Stddev	.0304	.0190
%RSD	17.58	11.42

#1	.1373	.1870
#2	.2118	.1412
#3	.1483	.1622
#4	.1899	.1829
#5	.1780	.1569

2187

Sample Name: s98t002992_d Run Time: 03/01/99 21:40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	142.52
Stddev	1.37
%RSD	.96248

#1	140.84
#2	141.63
#3	143.86
#4	143.96
#5	142.31

2188

Method: ACTINIDE Sample Name: s98t002993

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:44 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1803	.1746	3.354	285.2	.1614
Stddev	.0209	.0118	.094	1.8	.0213
%RSD	11.62	6.740	2.796	.6428	13.19

#1	.1828	.1687	3.277	283.1	.1242
#2	.1864	.1864	3.393	287.6	.1746
#3	.2080	.1584	3.300	286.5	.1631
#4	.1503	.1738	3.502	284.9	.1715
#5	.1737	.1854	3.300	284.0	.1737

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5906	.4822	.1812	.1779	.1727
Stddev	.0329	.0306	.0147	.0171	.0274
%RSD	5.575	6.355	8.087	9.632	15.89

#1	.5836	.5109	.1805	.1828	.1664
#2	.6275	.5025	.2029	.1581	.1416
#3	.6147	.4397	.1868	.1891	.2128
#4	.5426	.4604	.1668	.1973	.1574
#5	.5845	.4977	.1690	.1620	.1854

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1845	.1878
Stddev	.0242	.0150
%RSD	13.14	7.983

#1	.1523	.1781
#2	.1675	.1817
#3	.2009	.2128
#4	.2114	.1903
#5	.1901	.1761

Sample Name: s98t002993   Run Time: 03/01/99 21:44

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	141.69
Stddev	.52
%RSD	.36503
#1	142.22
#2	141.29
#3	141.00
#4	141.91
#5	142.00

Method: ACTINIDE

Sample Name: s98t002993_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 21:49 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1722	.1684	3.509	283.7	.1742
Stddev	.0270	.0247	.134	3.4	.0181
%RSD	15.66	14.66	3.830	1.188	10.38

#1	.1732	.1876	3.609	285.6	.1973
#2	.1711	.1450	3.327	277.9	.1545
#3	.2157	.1989	3.599	284.0	.1582
#4	.1569	.1665	3.403	285.0	.1855
#5	.1442	.1442	3.608	286.1	.1755

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5645	.4431	.1894	.1574	.1918
Stddev	.0444	.0353	.0190	.0185	.0194
%RSD	7.868	7.977	10.04	11.77	10.10

#1	.5485	.4402	.1612	.1852	.1949
#2	.5133	.4325	.1972	.1473	.2186
#3	.5679	.5008	.2133	.1677	.1917
#4	.6349	.4042	.1902	.1427	.1641
#5	.5576	.4375	.1851	.1442	.1899

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1695	.1799
Stddev	.0269	.0198
%RSD	15.89	11.02

#1	.1901	.1997
#2	.1402	.1806
#3	.1989	.1606
#4	.1427	.1593
#5	.1755	.1995

Sample Name: s98t002993_d Run Time: 03/01/99 21:49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	139.36
Stddev	.82
%RSD	.58551
#1	138.56
#2	140.26
#3	139.10
#4	140.18
#5	138.68

Method: ACTINIDE Sample Name: s98t002993_a Operator: bjg
 Comment: 201 df
 Run Time: 03/01/99 21:53 Type: Unk Mode: IR Corr. Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1750	.3185	1850.	286.6	2.331
Stddev	.0201	.0178	38.	3.5	.092
%RSD	11.47	5.586	2.053	1.208	3.931
#1	.1549	.2958	1788.	280.9	2.178
#2	.1600	.3223	1844.	289.4	2.337
#3	.1753	.3050	1883.	289.4	2.389
#4	.1790	.3386	1875.	287.3	2.414
#5	.2061	.3307	1863.	286.2	2.336

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2228.	1995.	53.96	.3666	.1671
Stddev	36.	56.	1.56	.0321	.0354
%RSD	1.597	2.787	2.896	8.760	21.20
#1	2165.	1900.	51.22	.3756	.1925
#2	2245.	1997.	54.17	.3223	.1647
#3	2235.	2007.	54.67	.4010	.2089
#4	2245.	2038.	55.00	.3458	.1185
#5	2251.	2032.	54.76	.3882	.1510

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2218.	.2253
Stddev	38.	.0325
%RSD	1.708	14.43
#1	2160.	.2042
#2	2216.	.1838
#3	2212.	.2497
#4	2258.	.2636
#5	2245.	.2253

Sample Name: s98t002993_a Run Time: 03/01/99 21:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	139.47
Stddev	1.56
%RSD	1.1160

#1	142.00
#2	139.62
#3	138.81
#4	137.83
#5	139.10

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/01/99 22:01 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2005	.2567	969.9	15.68	1.368
Stddev	.0176	.0317	11.2	.27	.031
%RSD	8.784	12.37	1.153	1.712	2.295

#1	.1708	.2340	953.4	15.46	1.343
#2	.2173	.2340	980.0	16.13	1.406
#3	.2068	.2549	971.5	15.69	1.387
#4	.2067	.2495	965.0	15.63	1.373
#5	.2009	.3109	979.9	15.50	1.330

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1146.	1026.	27.96	.2891	.1942
Stddev	8.	11.	.43	.0314	.0100
%RSD	.7022	1.094	1.542	10.88	5.131

#1	1148.	1010.	27.67	.2504	.1942
#2	1152.	1033.	27.65	.2937	.1815
#3	1153.	1040.	28.70	.3318	.2020
#4	1134.	1022.	27.96	.3018	.1877
#5	1140.	1024.	27.81	.2679	.2057

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1149.	.2039
Stddev	11.	.0207
%RSD	.9916	10.16

#1	1144.	.1778
#2	1158.	.2220
#3	1163.	.2140
#4	1135.	.1853
#5	1144.	.2201

Sample Name: ccv Run Time: 03/01/99 22:01

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	140.07
Stddev	1.46
%RSD	1.0421

#1	142.46
#2	139.61
#3	138.63
#4	140.28
#5	139.36

HNF-1663 REV.0

Method: ACTINIDE

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 22:09 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1745	.1721	.4287	.2393	.1611
Stddev	.0227	.0223	.0303	.0161	.0192
%RSD	13.02	12.94	7.075	6.710	11.90

#1	.1798	.1990	.4172	.2566	.1439
#2	.1581	.1916	.4575	.2299	.1748
#3	.2005	.1671	.4631	.2554	.1862
#4	.1892	.1533	.3927	.2203	.1580
#5	.1449	.1498	.4131	.2343	.1425

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1784	.5604	.1784	.1780	.1707
Stddev	.0106	.0802	.0088	.0276	.0348
%RSD	5.952	14.32	4.906	15.51	20.40

#1	.1702	.5323	.1798	.1870	.2158
#2	.1940	.5413	.1796	.2036	.1581
#3	.1671	.6707	.1719	.1766	.1766
#4	.1820	.6010	.1916	.1317	.1820
#5	.1787	.4565	.1691	.1908	.1208

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1889	.1985
Stddev	.0278	.0213
%RSD	14.70	10.71

#1	.2302	.1798
#2	.1581	.1892
#3	.1814	.2315
#4	.2011	.1844
#5	.1739	.2077

2197

Sample Name: ccb Run Time: 03/01/99 22:09

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	139.01
Stddev	.61
%RSD	.44141

#1	139.01
#2	139.17
#3	139.65
#4	139.21
#5	138.00

LABCORE Completed Worklist Report for Worklist# 27270

Analyst: bjg Instrument: ICPMS1

Book#: 40855, 990212-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) ISO_U

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000714	7.57e-04	106.022 %	Recovery
1	ICV	0	QMSU-QC U238	QC	0.100	1.04e-01	104.000 %	Recovery
2	ICB	0	QMSU-QC U233	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5		ug/mL
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5		ug/mL
3	STD-PREP	0	QMSU-A1 U238-A1	SOLID	10.0	1.07e+01	107.000 %	Recovery
4	BLNK-PREP	0	QMSU-A1 U233-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	QMSU-A1 U234-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	QMSU-A1 U235-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	QMSU-A1 U236-A1	SOLID	1	<1.60e-4		ug/g
4	BLNK-PREP	0	QMSU-A1 U238-A1	SOLID	1	<1.20e-4		ug/g
5	SAMPLE	S98T002990	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.678e-01	0.468	ug/g
5	SAMPLE	S98T002990	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.678e-01	0.468	ug/g
5	SAMPLE	S98T002990	0 A QMSU-A1 U235-A1	SOLID	N/A	1.244e+00	0.468	ug/g
5	SAMPLE	S98T002990	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.237e-01	0.624	ug/g
5	SAMPLE	S98T002990	0 A QMSU-A1 U238-A1	SOLID	N/A	1.766e+02	0.468	ug/g
6	DUP	S98T002990	0 A QMSU-A1 U233-A1	SOLID	<4.68e-1	<4.80e-1		RPD
6	DUP	S98T002990	0 A QMSU-A1 U234-A1	SOLID	<4.68e-1	<4.80e-1		RPD
6	DUP	S98T002990	0 A QMSU-A1 U235-A1	SOLID	1.24e+00	8.56e-01	36.641	RPD
6	DUP	S98T002990	0 A QMSU-A1 U236-A1	SOLID	<6.24e-1	<6.40e-1		RPD
6	DUP	S98T002990	0 A QMSU-A1 U238-A1	SOLID	1.77e+02	1.25e+02	34.437	RPD
7	SPK	S98T002990	0 A QMSU-A1 U235-A1	SOLID	4.00	3.96e+00	99.000 %	Recovery
7	SPK	S98T002990	0 A QMSU-A1 U238-A1	SOLID	2000.00	1.86e+03	93.000 %	Recovery
8	SPK-POST	S98T002990	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.59e-04	111.814 %	Recovery
8	SPK-POST	S98T002990	0 A QMSU-A1 U238-A1	SOLID	0.020	2.18e-02	109.000 %	Recovery
9	SAMPLE	S98T002991	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.782e-01	0.478	ug/g
9	SAMPLE	S98T002991	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.782e-01	0.478	ug/g
9	SAMPLE	S98T002991	0 A QMSU-A1 U235-A1	SOLID	N/A	1.598e+00	0.478	ug/g
9	SAMPLE	S98T002991	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.376e-01	0.638	ug/g
9	SAMPLE	S98T002991	0 A QMSU-A1 U238-A1	SOLID	N/A	2.443e+02	0.478	ug/g
10	DUP	S98T002991	0 A QMSU-A1 U233-A1	SOLID	<4.78e-1	<4.90e-1		RPD
10	DUP	S98T002991	0 A QMSU-A1 U234-A1	SOLID	<4.78e-1	<4.90e-1		RPD
10	DUP	S98T002991	0 A QMSU-A1 U235-A1	SOLID	1.60e+00	1.63e+00	1.858	RPD
10	DUP	S98T002991	0 A QMSU-A1 U236-A1	SOLID	<6.38e-1	<6.53e-1		RPD
10	DUP	S98T002991	0 A QMSU-A1 U238-A1	SOLID	2.44e+02	2.34e+02	4.184	RPD
11	SPK	S98T002991	0 A QMSU-A1 U235-A1	SOLID	4.00	4.03e+00	100.750 %	Recovery
11	SPK	S98T002991	0 A QMSU-A1 U238-A1	SOLID	2000.00	1.98e+03	99.000 %	Recovery
12	SPK-POST	S98T002991	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.61e-04	113.221 %	Recovery
12	SPK-POST	S98T002991	0 A QMSU-A1 U238-A1	SOLID	0.020	2.16e-02	108.000 %	Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27270

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
13	CCV	0	EMSU-QC U235	QC	0.000142	1.60e-04	112.676 % Recovery	
13	CCV	0	EMSU-QC U238	QC	0.02000	2.23e-02	111.500 % Recovery	
14	CCB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
14	CCB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
14	CCB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
14	CCB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
14	CCB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	
15	SAMPLE	S98T002992	0 A EMSU-A1 U233-A1	SOLID	N/A	< 4.713e-01	0.471 ug/g	
15	SAMPLE	S98T002992	0 A EMSU-A1 U234-A1	SOLID	N/A	< 4.713e-01	0.471 ug/g	
15	SAMPLE	S98T002992	0 A EMSU-A1 U235-A1	SOLID	N/A	1.233e+00	0.471 ug/g	
15	SAMPLE	S98T002992	0 A EMSU-A1 U236-A1	SOLID	N/A	< 6.284e-01	0.628 ug/g	
15	SAMPLE	S98T002992	0 A EMSU-A1 U238-A1	SOLID	N/A	1.767e+02	0.471 ug/g	
16	DUP	S98T002992	0 A EMSU-A1 U233-A1	SOLID	<4.71e-1	<4.78e-1	RPD	
16	DUP	S98T002992	0 A EMSU-A1 U234-A1	SOLID	<4.71e-1	<4.78e-1	RPD	
16	DUP	S98T002992	0 A EMSU-A1 U235-A1	SOLID	1.23e+00	1.04e+00	16.740 RPD	
16	DUP	S98T002992	0 A EMSU-A1 U236-A1	SOLID	<6.28e-1	<6.38e-1	RPD	
16	DUP	S98T002992	0 A EMSU-A1 U238-A1	SOLID	1.77e+02	1.61e+02	9.467 RPD	
17	SAMPLE	S98T002993	0 A EMSU-A1 U233-A1	SOLID	N/A	< 4.811e-01	0.481 ug/g	
17	SAMPLE	S98T002993	0 A EMSU-A1 U234-A1	SOLID	N/A	< 4.811e-01	0.481 ug/g	
17	SAMPLE	S98T002993	0 A EMSU-A1 U235-A1	SOLID	N/A	8.739e-01	0.481 ug/g	
17	SAMPLE	S98T002993	0 A EMSU-A1 U236-A1	SOLID	N/A	< 6.414e-01	0.641 ug/g	
17	SAMPLE	S98T002993	0 A EMSU-A1 U238-A1	SOLID	N/A	1.179e+02	0.481 ug/g	
18	DUP	S98T002993	0 A EMSU-A1 U233-A1	SOLID	<4.81e-1	<4.95e-1	RPD	
18	DUP	S98T002993	0 A EMSU-A1 U234-A1	SOLID	<4.81e-1	<4.95e-1	RPD	
18	DUP	S98T002993	0 A EMSU-A1 U235-A1	SOLID	8.74e-01	9.98e-01	13.248 RPD	
18	DUP	S98T002993	0 A EMSU-A1 U236-A1	SOLID	<6.41e-1	<6.60e-1	RPD	
18	DUP	S98T002993	0 A EMSU-A1 U238-A1	SOLID	1.18e+02	1.23e+02	4.149 RPD	
19	SPK	S98T002993	0 A EMSU-A1 U235-A1	SOLID	4.000	4.31e+00	107.750 % Recovery	
19	SPK	S98T002993	0 A EMSU-A1 U238-A1	SOLID	2000.00	2.16e+03	108.000 % Recovery	
20	SPK-POST	S98T002993	0 A EMSU-A1 U235-A1	SOLID	0.0001422	1.62e-04	113.924 % Recovery	
20	SPK-POST	S98T002993	0 A EMSU-A1 U238-A1	SOLID	0.020	2.30e-02	115.000 % Recovery	
21	CCV	0	EMSU-QC U235	QC	0.0001422	1.63e-04	114.627 % Recovery	
21	CCV	0	EMSU-QC U238	QC	0.020	2.24e-02	112.000 % Recovery	
22	CCB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
22	CCB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
22	CCB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
22	CCB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
22	CCB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	

Final page for worklist# 27270

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Paul M. Pang 2/18/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27270

Analyst: BTC Instrument: ICPMS1 _____ Book# 40855 59012-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) ISO_U

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV			@MSU-QC	QC		
2	ICB			@MSU-QC	QC		
3	STD-PREP			@MSU-A1	SOLID		
4	BLNK-PREP			@MSU-A1	SOLID		
5	SAMPLE	S98T002990	0	A	@MSU-A1	SOLID	98000509 BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
6	DUP	S98T002990	0	A	@MSU-A1	SOLID	
7	SPK	S98T002990	0	A	@MSU-A1	SOLID	
8	SPK-POST	S98T002990	0	A	@MSU-A1	SOLID	
9	SAMPLE	S98T002991	0	A	@MSU-A1	SOLID	98000509 BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
10	DUP	S98T002991	0	A	@MSU-A1	SOLID	
11	SPK	S98T002991	0	A	@MSU-A1	SOLID	
12	SPK-POST	S98T002991	0	A	@MSU-A1	SOLID	
13	CCV			@MSU-QC	QC		
14	CCB			@MSU-QC	QC		
15	SAMPLE	S98T002992	0	A	@MSU-A1	SOLID	98000509 BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
16	DUP	S98T002992	0	A	@MSU-A1	SOLID	
17	SAMPLE	S98T002993	0	A	@MSU-A1	SOLID	98000509 BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					

Data Entry Comments:

Uploaded & Validated by:
S. M. Page
2/18/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code

2201

LABCORE Data Entry Template for Worklist# 27270

S Type	Sample#	R	A	Test	Matrix	Group#	Project
18 DUP	S98T002993	0	A	@MSU-A1	SOLID		
19 SPK	S98T002993	0	A	@MSU-A1	SOLID		
20 SPK-POST	S98T002993	0	A	@MSU-A1	SOLID		
21 CCV				@MSU-QC	QC		
22 CCB				@MSU-QC	QC		

Final page for worklist # 27270

Paul Guder 2-17-98
Signature Date

Signature Date

prep STD .05-10 ~ 1
prep Bck 1-9 ~ 1
S98T002990 .05-10 ~ 1

DF
201
10
201

90-D
90-S
90-A
91
91-D
91-S
91-A
92
92-D
93
93-D
93-S
93-A

Data Entry Comments:

POST SPK (A) = .020 ~ 1 of 10 ppm U STD

SCANS = 982990, 982991, 982992, 982993

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 02/17/99 13:20 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	25500.
Stddev	301.
%RSD	1.181

#1	25800.
#2	25170.
#3	25220.
#4	25440.
#5	25410.
#6	25780.
#7	25970.
#8	25620.
#9	25040.
#10	25570.

2-17-99

BY-105 AND BEST

SFTW 2990-93

.05-.01 201 A.F.

FILE # 402176.TXT

W.L.# 27220

SCANS = 982990

882981

982992

982993

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2203 TO 2232.

HNF-1663 REV. 0

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/17/99 13:21 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.046	5.638	4.472	4.737	5.779
Stddev	.630	.886	.452	.671	.374
%RSD	12.48	15.72	10.11	14.17	6.471
#1	5.927	5.806	4.597	5.686	5.565
#2	5.435	4.371	4.608	4.253	6.144
#3	4.877	5.471	3.687	4.163	5.233
#4	4.384	5.688	4.858	4.384	5.925
#5	4.609	6.854	4.609	5.200	6.027

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.2039
Stddev	.0411
%RSD	.97720

#1	4.1333
#2	4.2320
#3	4.2037
#4	4.2194
#5	4.2309

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	507.3996	5.0464	1.0000000	02/17/99 13:23:15

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.0464
100ppb U	0.711	0.711	0	0	365.81

U [233]

0
(S) IR⁴¹⁰⁰¹⁰ Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	506.5672	5.6383	1.0000000	02/17/99 13:23:15

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.6383
100ppb U	0.711	0.711	0	0	365.81

U [234]

0
(S) IR⁴¹⁰⁰¹⁰ Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	508.2078	4.4718	1.0000000	02/17/99 13:23:15

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4718
100ppb U	0.711	0.711	0	0	365.81

U [235]

0
(S) IR⁴¹⁰⁰¹⁰ Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	507.8345	4.7372	1.0000000	02/17/99 13:23:15

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.7372
100ppb U	0.711	0.711	0	0	365.81

U [236]

0
(S) IR⁴¹⁰⁰¹⁰ Concentration

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_{238}	469.4287	5.7787	1.0000000	02/17/99 13:23:15

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.7787
100ppb U	99.3	99.3	0	0	46620

U [238]

0

51290MTU
(S)IR

Concentration

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/17/99 13:24 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	329.5	329.5	329.5	329.5	43240.
Stdddev	11.2	11.2	11.2	11.2	776.
%RSD	3.391	3.391	3.391	3.391	1.795
#1	334.5	334.5	334.5	334.5	43530.
#2	332.5	332.5	332.5	332.5	43710.
#3	341.6	341.6	341.6	341.6	43980.
#4	327.0	327.0	327.0	327.0	42930.
#5	311.9	311.9	311.9	311.9	42040.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3061
Stdddev	.1266
%RSD	2.9396

#1	4.2181
#2	4.2566
#3	4.2766
#4	4.2497
#5	4.5294

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	456.3307	5.0464	1.0000000	02/17/99 13:25:58

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.0464
100ppb U	0.711	0.711	0	0	329.5

U [233]

0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	455.4983	5.6383	1.0000000	02/17/99 13:25:58

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.6383
100ppb U	0.711	0.711	0	0	329.5

U [234]

0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	457.1389	4.4718	1.0000000	02/17/99 13:25:58

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4718
100ppb U	0.711	0.711	0	0	329.5

U [235]

0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	456.7656	4.7372	1.0000000	02/17/99 13:25:58

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.7372
100ppb U	0.711	0.711	0	0	329.5

U [236]

0
(S) IR³⁷⁰⁰¹⁰

Concentration

HNF-1663 REV: 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	435.3836	5.7787	1.0000000	02/17/99 13:25:58

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.7787
100ppb U	99.3	99.3	0	0	43239

U [238]

0

(S)IR

Concentration

HNF-1663 REV. 0

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 02/17/99 13:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0008	-.0025	.0001	-.0011	.0000
Stddev	.0024	.0009	.0007	.0015	.0020
%RSD	302.8	37.47	627.0	135.0	4828.
#1	.0007	-.0021	-.0001	.0011	-.0033
#2	-.0013	-.0036	.0012	-.0016	.0012
#3	-.0022	-.0022	-.0004	-.0012	-.0002
#4	-.0037	-.0013	.0000	-.0008	.0011
#5	.0025	-.0033	-.0003	-.0031	.0014

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3808
Stddev	.0560
%RSD	1.2781

#1	4.3867
#2	4.3688
#3	4.3084
#4	4.4650
#5	4.3752

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/17/99 13:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0005	.0058	.7568	.0126	103.7
Stddev	.0017	.0035	.0284	.0031	1.4
%RSD	354.0	59.30	3.756	24.39	1.315
#1	.0008	.0056	.7118	.0118	101.4
#2	.0020	.0012	.7864	.0167	104.7
#3	.0021	.0071	.7523	.0104	104.8
#4	-.0006	.0046	.7592	.0093	103.6
#5	-.0019	.0107	.7742	.0147	104.1

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3493
Stddev	.0721
%RSD	1.6574

#1	4.4534
#2	4.2860
#3	4.3317
#4	4.3895
#5	4.2859

HNF-1663 REV. 0

Method: ISOU

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 02/17/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0032	.0005	-.0016	-.0008
Stddev	.0015	.0004	.0010	.0007	.0030
%RSD	170.3	11.94	181.2	44.75	381.3
#1	-.0019	-.0027	-.0001	-.0004	.0020
#2	-.0028	-.0036	-.0006	-.0016	.0024
#3	-.0004	-.0031	.0013	-.0018	-.0042
#4	.0006	-.0033	.0018	-.0023	-.0032
#5	.0003	-.0035	.0003	-.0021	-.0009

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3495
Stddev	.3513
%RSD	8.0758

#1	3.8438
#2	4.3910
#3	4.8350
#4	4.3311
#5	4.3463

Method: ISOU Sample Name: std-prep

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 13:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0007	.1112	.0058	53.29
Stddev	.0018	.0013	.0071	.0016	2.17
%RSD	418.1	173.3	6.404	27.26	4.073
#1	-.0002	-.0004	.1208	.0076	55.31
#2	-.0028	-.0008	.1101	.0038	53.21
#3	-.0015	-.0028	.1018	.0072	49.78
#4	.0009	-.0001	.1084	.0058	53.32
#5	.0015	.0005	.1149	.0049	54.84

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.1748
Stddev	.1393
%RSD	3.3366

#1	4.0194
#2	4.2590
#3	4.3702
#4	4.1277
#5	4.0975

Method: ISOU

Sample Name: blnk-prep

Operator: bjg

Comment: 10 df

Run Time: 02/17/99 13:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0024	-.0023	.0022	.0011	.0025
Stddev	.0028	.0012	.0028	.0022	.0018
%RSD	117.9	53.12	124.1	198.1	70.01
#1	.0028	-.0007	-.0004	.0009	.0005
#2	.0044	-.0017	.0060	-.0013	.0039
#3	.0028	-.0041	.0038	.0015	.0034
#4	-.0024	-.0025	.0023	.0046	.0041
#5	.0044	-.0025	-.0005	-.0002	.0007

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.4775
Stddev	.1145
%RSD	3.2934

#1	3.4911
#2	3.6125
#3	3.3091
#4	3.4348
#5	3.5399

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002990

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	-.0035	.0319	-.0002	4.531
Stddev	.0025	.0017	.0022	.0026	.210
%RSD	1070.	46.97	6.982	1325.	4.635
#1	-.0016	-.0020	.0311	.0000	4.412
#2	.0000	-.0041	.0286	-.0016	4.446
#3	.0046	-.0019	.0338	.0041	4.632
#4	-.0007	-.0059	.0340	-.0030	4.318
#5	-.0012	-.0037	.0323	-.0005	4.846

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.6521
Stddev	.1582
%RSD	4.3303

#1	3.6911
#2	3.8503
#3	3.5671
#4	3.7184
#5	3.4338

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002990_d

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0022	-.0014	.0214	-.0009	3.133
Stddev	.0015	.0015	.0028	.0023	.131
%RSD	67.85	105.0	13.03	241.6	4.171
#1	-.0026	.0010	.0188	.0017	3.158
#2	-.0025	-.0023	.0189	-.0041	2.973
#3	.0004	-.0012	.0253	.0008	3.209
#4	-.0030	-.0017	.0208	-.0020	3.031
#5	-.0031	-.0030	.0230	-.0010	3.294

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.3301
Stddev	.1210
%RSD	3.6332

#1	3.3674
#2	3.4704
#3	3.2441
#4	3.3980
#5	3.1705

Method: ISOU

Sample Name: s98t002990_s

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0023	-.0015	.1366	.0057	53.32
Stddev	.0017	.0006	.0186	.0022	1.47
%RSD	73.20	43.35	13.60	38.58	2.765
#1	.0010	-.0024	.1616	.0078	54.00
#2	.0005	-.0012	.1440	.0032	52.59
#3	.0020	-.0014	.1397	.0041	52.77
#4	.0034	-.0017	.1132	.0081	51.71
#5	.0045	-.0007	.1246	.0052	55.51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.1218
Stddev	.0460
%RSD	1.4734

#1	3.0766
#2	3.1366
#3	3.1098
#4	3.1938
#5	3.0923

Method: ISOU

Sample Name: s98t002990_a

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0021	.0008	.1910	.0044	26.37
Stddev	.0028	.0032	.0124	.0027	.85
%RSD	132.0	391.6	6.506	60.80	3.231
#1	-.0025	-.0007	.1915	.0032	26.54
#2	.0016	.0065	.1711	.0043	25.76
#3	.0037	-.0015	.1896	.0039	26.88
#4	.0048	-.0008	.2013	.0016	25.28
#5	.0032	.0007	.2017	.0087	27.40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	2.7716
Stddev	.0780
%RSD	2.8159

#1	2.8149
#2	2.6787
#3	2.8300
#4	2.8393
#5	2.6951

HNF-1663 REV.0

Method: ISOU

Sample Name: s98t002991

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0040	-.0026	.0401	.0024	6.130
Stddev	.0021	.0008	.0060	.0022	.430
%RSD	53.55	30.17	14.83	90.43	7.022
#1	.0034	-.0034	.0461	-.0011	6.815
#2	.0030	-.0023	.0311	.0017	6.179
#3	.0063	-.0015	.0375	.0039	6.130
#4	.0060	-.0032	.0422	.0040	5.742
#5	.0012	-.0027	.0439	.0035	5.785

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.2804
Stddev	.0832
%RSD	2.5371

#1	3.1919
#2	3.2642
#3	3.2147
#4	3.3459
#5	3.3853

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002991_d

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:25 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_ [233]	U_ [234]	U_ [235]	U_ [236]	U_ [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0017	-.0031	.0400	-.0011	5.737
Stddev	.0021	.0020	.0056	.0008	.252
%RSD	123.0	62.29	13.98	77.54	4.388
#1	.0033	-.0043	.0450	-.0011	5.437
#2	.0028	-.0056	.0341	-.0006	5.702
#3	-.0009	-.0022	.0410	-.0003	5.628
#4	.0036	-.0005	.0343	-.0025	5.797
#5	-.0002	-.0031	.0456	-.0011	6.120

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	3.5622
Stddev	.0869
%RSD	2.4387

#1	3.6521
#2	3.5670
#3	3.5748
#4	3.5982
#5	3.4186

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002991_a

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:28 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0038	-.0018	.2007	.0028	27.70
Stddev	.0015	.0017	.0142	.0021	1.92
%RSD	39.16	97.88	7.050	76.34	6.943
#1	.0035	-.0024	.2152	.0020	28.83
#2	.0029	-.0028	.2017	.0003	26.96
#3	.0064	.0008	.1939	.0031	28.08
#4	.0028	-.0035	.1805	.0023	24.81
#5	.0037	-.0009	.2123	.0061	29.82

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.7683
Stddev	.1397
%RSD	3.7073

#1	3.6212
#2	3.7810
#3	3.8236
#4	3.9660
#5	3.6496

HNF-1663 REV.0

Method: ISOU

Sample Name: s98t002991_s

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0014	-.0018	.1486	.0063	58.88
Stddev	.0014	.0016	.0087	.0019	2.54
%RSD	97.58	93.05	5.849	29.81	4.309
#1	.0021	-.0033	.1453	.0042	54.54
#2	-.0007	-.0017	.1372	.0065	60.93
#3	.0013	-.0018	.1584	.0067	59.39
#4	.0031	.0009	.1562	.0090	60.43
#5	.0014	-.0029	.1462	.0050	59.11

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.6984
Stddev	.1830
%RSD	4.9479

#1	3.9923
#2	3.7069
#3	3.5370
#4	3.5520
#5	3.7035

HNF-1663 REV. 0

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: 20 ppb u std

Run Time: 02/17/99 14:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	-.0014	.1601	.0045	22.27
Stddev	.0011	.0007	.0132	.0027	1.17
%RSD	.235.3	50.07	8.239	59.80	5.240
#1	.0005	-.0005	.1549	.0044	21.02
#2	-.0014	-.0014	.1637	.0025	21.59
#3	-.0012	-.0012	.1653	.0066	23.81
#4	.0010	-.0024	.1759	.0078	23.15
#5	-.0013	-.0016	.1406	.0013	21.78

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	3.3178
Stddev	.1020
%RSD	3.0740

#1	3.4068
#2	3.3978
#3	3.2293
#4	3.1875
#5	3.3675

Method: ISOU

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/17/99 14:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0020	-.0025	.0004	-.0012	-.0023
Stddev	.0010	.0012	.0018	.0014	.0019
%RSD	51.71	47.93	428.3	116.5	82.88

#1	-.0023	-.0033	-.0014	-.0016	.0003
#2	-.0017	-.0037	-.0002	-.0007	-.0049
#3	-.0026	-.0030	-.0001	-.0026	-.0024
#4	-.0030	-.0007	.0002	-.0020	-.0017
#5	-.0003	-.0020	.0034	.0010	-.0027

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.3814
Stddev	.0636
%RSD	1.8814

#1	3.3736
#2	3.2943
#3	3.3820
#4	3.3831
#5	3.4739

Method: ISOU

Sample Name: s98t002992

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_ [233]	U_ [234]	U_ [235]	U_ [236]	U_ [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0011	-.0034	.0314	.0003	4.500
Stddev	.0010	.0013	.0017	.0023	.126
%RSD	93.76	37.54	5.384	675.0	2.807
#1	.0021	-.0022	.0295	.0037	4.366
#2	-.0002	-.0043	.0309	-.0004	4.486
#3	.0015	-.0050	.0306	-.0017	4.429
#4	.0020	-.0033	.0319	.0016	4.519
#5	.0002	-.0022	.0340	-.0015	4.701

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	3.2813
Stddev	.1046
%RSD	3.1881

#1	3.3405
#2	3.4159
#3	3.1439
#4	3.2819
#5	3.2241

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002992_d

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	-.0021	.0262	.0012	4.029
Stddev	.0023	.0014	.0041	.0024	.139
%RSD	518.2	68.29	15.76	201.4	3.441
#1	.0018	-.0024	.0252	.0032	4.235
#2	-.0002	-.0035	.0275	.0015	4.023
#3	.0000	-.0033	.0227	.0003	3.870
#4	-.0044	-.0011	.0228	.0033	4.073
#5	.0005	-.0002	.0326	-.0025	3.945

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.1823
Stddev	.1390
%RSD	4.3692

#1	2.9736
#2	3.2232
#3	3.2683
#4	3.1198
#5	3.3266

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002993

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 14:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0036	.0218	-.0002	2.940
Stddev	.0025	.0013	.0034	.0019	.160
%RSD	738.2	36.25	15.74	876.3	5.425
#1	.0006	-.0041	.0181	-.0024	2.681
#2	-.0031	-.0017	.0184	-.0004	3.067
#3	.0027	-.0043	.0257	-.0005	2.900
#4	-.0027	-.0051	.0228	.0028	2.996
#5	.0009	-.0029	.0241	-.0006	3.057

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.2522
Stddev	.1944
%RSD	5.9767

#1	3.5744
#2	3.2956
#3	3.1113
#4	3.1516
#5	3.1279

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002993_d

Operator: bjg

Comment: 201 df

Run Time: 02/17/99

15:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0007	-.0021	.0242	.0025	2.990
Stddev	.0020	.0027	.0042	.0019	.169
%RSD	268.3	126.0	17.31	74.06	5.651
#1	-.0037	.0006	.0264	.0052	3.058
#2	.0002	-.0033	.0293	.0036	2.926
#3	-.0019	-.0061	.0184	.0014	3.247
#4	.0010	-.0010	.0249	.0010	2.907
#5	.0006	-.0007	.0220	.0013	2.810

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.0084
Stddev	.0834
%RSD	2.7731

#1	2.9595
#2	2.9070
#3	2.9870
#4	3.0913
#5	3.0969

Analysis Report

HNF-1663 REV. 0

Method: ISOU

Sample Name: S98t002993_a

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 15:03 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0008	-.0015	.1840	.0020	25.96
Stddev	.0026	.0022	.0123	.0026	1.31
%RSD	345.7	143.4	6.692	129.5	5.043
#1	-.0008	-.0014	.1708	.0010	24.82
#2	.0054	-.0027	.1903	-.0018	27.41
#3	.0000	-.0024	.1932	.0029	26.00
#4	.0002	.0022	.1705	.0052	24.49
#5	-.0009	-.0034	.1952	.0028	27.08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	2.9480
Stddev	.0788
%RSD	2.6730

#1	2.9922
#2	2.9359
#3	2.9800
#4	3.0150
#5	2.8168

Analysis Report

02/17/99 15:07:40

page 1

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002993_s

Operator: bjg

Comment: 201 df

Run Time: 02/17/99 15:06 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem.	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0008	-.0019	.1329	.0068	58.24
Stddev	.0016	.0030	.0165	.0039	2.86
%RSD	192.9	160.0	12.42	58.05	4.913
#1	-.0033	-.0046	.1448	.0044	56.56
#2	.0003	-.0014	.1237	.0123	57.01
#3	-.0015	-.0024	.1141	.0018	55.09
#4	.0007	-.0039	.1547	.0077	61.57
#5	-.0004	.0030	.1273	.0075	60.96

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	2.7301
Stddev	.1356
%RSD	4.9685

#1	2.8085
#2	2.8029
#3	2.8695
#4	2.5995
#5	2.5699

Analysis Report

02/17/99 15:10:51

page 1

Method: ISOU
 Comment: 20 ppb u
 Run Time: 02/17/99 15:09 Type: Unk

HNF-1663 REV.0

Sample Name: ccv

Operator: bjg

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	.0000	.1627	.0015	22.36
Stddev	.0018	.002	.0175	.0019	2.20
%RSD	420.2	45980.	10.72	126.5	9.857
#1	-.0006	.0026	.1906	.0029	24.54
#2	.0024	-.0026	.1507	.0023	21.94
#3	-.0001	-.0006	.1539	.0006	23.95
#4	-.0014	.0000	.1691	.0031	22.44
#5	-.0024	.0006	.1493	-.0014	18.91

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	2.9185
Stddev	.2425
%RSD	8.3096

#1	2.6315
#2	2.9305
#3	2.8939
#4	2.8366
#5	3.2999

analysis Report

02/17/99 15:15:01

page 1

Method: ISOU
 Comment: 2% hno3
 Run Time: 02/17/99 15:13 Type: Unk

HNF-1663 REV.0
 Sample Name: ccb

Operator: bjg

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0018	-.0028	-.0005	-.0020	-.0051
Stddev	.0012	.0017	.0017	.0022	.0012
%RSD	67.85	62.20	356.8	111.2	23.97
#1	-.0027	-.0050	-.0018	-.0055	-.0071
#2	-.0005	-.0014	.0018	-.0020	-.0041
#3	-.0005	-.0025	.0001	-.0002	-.0043
#4	-.0030	-.0009	-.0001	.0000	-.0045
#5	-.0021	-.0041	-.0025	-.0024	-.0056

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	3.1851
Stddev	.1141
%RSD	3.5814

#1	3.1402
#2	3.0104
#3	3.2156
#4	3.2604
#5	3.2987

LABCORE Completed Worklist Report for Worklist# 27475

Analyst: bjg

Instrument: ICPMS1

Book#: 9859

Method: LA-506-101 Rev/Mod A0

Worklist Comment: ICP/MS BY-105 (DIRECT) Actinide

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	OMSACT	TH232	QC	0.010000	1.06e-02	106.000 % Recovery
1	ICV	0	OMSACT	NP237	QC	0.009840	1.41e-02	143.293 % Recovery
1	ICV	0	OMSACT	FU239	QC	0.010040	1.42e-02	141.434 % Recovery
1	ICV	0	OMSACT	241	QC	0.010260	1.63e-02	158.869 % Recovery
2	ICB	0	OMSACTA1	PA231-A1	LIQUID	1	<1.39e-5	ug/mL
2	ICB	0	OMSACTA1	FU239-A1	LIQUID	1	<1.29e-5	ug/mL
2	ICB	0	OMSACTA1	FU240-A1	LIQUID	1	<6.91e-6	ug/mL
2	ICB	0	OMSACTA1	FU242-A1	LIQUID	1	<1.04e-5	ug/mL
2	ICB	0	OMSACTA1	241-A1	LIQUID	1	<2.00e-5	ug/mL
2	ICB	0	OMSACTA1	243-A1	LIQUID	1	<1.48e-5	ug/mL
2	ICB	0	OMSACTA1	244-A1	LIQUID	1	<8.22e-6	ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	PA231-A1	LIQUID	N/A	< 5.646e-01 0.565 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	TH229-A1	LIQUID	N/A	< 1.034e+00 1.034 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	TH230-A1	LIQUID	N/A	< 6.742e-01 0.674 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	TH232-A1	LIQUID	N/A	< 5.553e-01 0.555 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	NP237-A1	LIQUID	N/A	< 1.058e+00 1.058 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	FU239-A1	LIQUID	N/A	< 5.218e-01 0.522 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	FU240-A1	LIQUID	N/A	< 2.800e-01 0.280 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	FU242-A1	LIQUID	N/A	< 4.195e-01 0.420 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	241-A1	LIQUID	N/A	< 8.104e-01 0.810 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	243-A1	LIQUID	N/A	< 5.993e-01 0.599 ug/mL
3	SAMPLE	S98T003012	0 A	OMSACTA1	244-A1	LIQUID	N/A	< 3.331e-01 0.333 ug/mL
4	DUP	S98T003012	0 A	OMSACTA1	PA231-A1	LIQUID	<5.65e-1	< 5.65e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	TH229-A1	LIQUID	<1.03e0	<1.03e0 RPD
4	DUP	S98T003012	0 A	OMSACTA1	TH230-A1	LIQUID	<6.74e-1	<6.74e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	TH232-A1	LIQUID	<5.55e-1	<5.55e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	NP237-A1	LIQUID	<1.06e0	<1.06e0 RPD
4	DUP	S98T003012	0 A	OMSACTA1	FU239-A1	LIQUID	<5.22e-1	<5.22e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	FU240-A1	LIQUID	<2.80e-1	<2.80e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	FU242-A1	LIQUID	<4.19e-1	<4.20e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	241-A1	LIQUID	<8.10e-1	<8.10e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	243-A1	LIQUID	<5.99e-1	<5.99e-1 RPD
4	DUP	S98T003012	0 A	OMSACTA1	244-A1	LIQUID	<3.33e-1	<3.33e-1 RPD
5	SPK-POST	S98T003012	0 A	OMSACTA1	TH232-A1	LIQUID	0.02000	1.78e-02 89.000 % Recovery
5	SPK-POST	S98T003012	0 A	OMSACTA1	NP237-A1	LIQUID	0.01976	2.56e-02 129.555 % Recovery
5	SPK-POST	S98T003012	0 A	OMSACTA1	FU239-A1	LIQUID	0.02008	2.58e-02 128.486 % Recovery
5	SPK-POST	S98T003012	0 A	OMSACTA1	241-A1	LIQUID	0.02052	3.03e-02 147.661 % Recovery
6	SAMPLE	S98T003013	0 A	OMSACTA1	PA231-A1	LIQUID	N/A	< 5.646e-01 0.565 ug/mL
6	SAMPLE	S98T003013	0 A	OMSACTA1	TH229-A1	LIQUID	N/A	< 1.034e+00 1.034 ug/mL
6	SAMPLE	S98T003013	0 A	OMSACTA1	TH230-A1	LIQUID	N/A	< 6.742e-01 0.674 ug/mL
6	SAMPLE	S98T003013	0 A	OMSACTA1	TH232-A1	LIQUID	N/A	< 5.553e-01 0.555 ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27475

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T003013	0 A	EMSACTA1 NP237-A1	LIQUID	N/A	< 1.058e+00	1.058	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 PU239-A1	LIQUID	N/A	< 5.218e-01	0.522	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 PU240-A1	LIQUID	N/A	< 2.800e-01	0.280	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 PU242-A1	LIQUID	N/A	< 4.195e-01	0.420	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 241-A1	LIQUID	N/A	< 8.104e-01	0.810	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 243-A1	LIQUID	N/A	< 5.993e-01	0.599	ug/mL
6 SAMPLE	S98T003013	0 A	EMSACTA1 244-A1	LIQUID	N/A	< 3.331e-01	0.333	ug/mL
7 DUP	S98T003013	0 A	EMSACTA1 PA231-A1	LIQUID	<5.65e-1	<5.65e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 TH229-A1	LIQUID	<1.03e0	<1.03e0		RPD
7 DUP	S98T003013	0 A	EMSACTA1 TH230-A1	LIQUID	<6.74e-1	<6.74e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 TH232-A1	LIQUID	<5.55e-1	<5.55e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 NP237-A1	LIQUID	<1.06e0	<1.06e0		RPD
7 DUP	S98T003013	0 A	EMSACTA1 PU239-A1	LIQUID	<5.22e-1	<5.22e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 PU240-A1	LIQUID	<2.80e-1	<2.80e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 PU242-A1	LIQUID	<4.19e-1	<4.20e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 241-A1	LIQUID	<8.10e-1	<8.10e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 243-A1	LIQUID	<5.99e-1	<5.99e-1		RPD
7 DUP	S98T003013	0 A	EMSACTA1 244-A1	LIQUID	<3.33e-1	<3.33e-1		RPD
8 CCV	0	0	EMSACT TH232	QC	0.01000	9.87e-03	98.700	% Recovery
8 CCV	0	0	EMSACT NP237	QC	0.00984	1.34e-02	136.179	% Recovery
8 CCV	0	0	EMSACT PU239	QC	0.01004	1.35e-02	134.462	% Recovery
8 CCV	0	0	EMSACT 241	QC	0.01026	1.60e-02	155.945	% Recovery
9 CCB	0	0	EMSACTA1 PA231-A1	LIQUID	1	<1.39e-5		ug/mL
9 CCB	0	0	EMSACTA1 TH229-A1	LIQUID	1	<2.55e-5		ug/mL
9 CCB	0	0	EMSACTA1 TH230-A1	LIQUID	1	<1.66e-5		ug/mL
9 CCB	0	0	EMSACTA1 TH232-A1	LIQUID	1	<1.37e-5		ug/mL
9 CCB	0	0	EMSACTA1 PU239-A1	LIQUID	1	<1.29e-5		ug/mL
9 CCB	0	0	EMSACTA1 PU240-A1	LIQUID	1	<6.91e-6		ug/mL
9 CCB	0	0	EMSACTA1 PU242-A1	LIQUID	1	<1.04e-5		ug/mL
9 CCB	0	0	EMSACTA1 241-A1	LIQUID	1	<2.00e-5		ug/mL
9 CCB	0	0	EMSACTA1 243-A1	LIQUID	1	<1.48e-5		ug/mL
9 CCB	0	0	EMSACTA1 244-A1	LIQUID	1	<8.22e-6		ug/mL

Final page for worklist# 27475

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. P... 01/06/99
Reviewer Signature _____ Date _____

12/09/98 11:14
A-0004-1

FILE # A-0006A-TXT

Page: 1

LABCORE Data Entry Template for Worklist# 27475

Analyst: B. G. Miller Instrument: ICPMS1 _____ Book# 9859Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) Actinide

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICB			@MSACTAL	LIQUID		
3 SAMPLE	S98T003012 0 A		@MSACTAL	LIQUID	98000510	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
4 DUP	S98T003012 0 A		@MSACTAL	LIQUID		
5 SPK-POST	S98T003012 0 A		@MSACTAL	LIQUID		
6 SAMPLE	S98T003013 0 A		@MSACTAL	LIQUID	98000510	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
7 DUP	S98T003013 0 A		@MSACTAL	LIQUID		
8 CCV			@MSACT	QC		
9 CCB			@MSACTAL	LIQUID		

Final page for worklist # 27475

B. G. Miller 1-6-99
 Analyst Signature Date

Analyst Signature Date

S98T003012 1-10-025-10
 3012-D 1-10-025-10
 3012-A 1-10-025-10
 3013 1-10-025-10
 3013-1 1-10-025-10

DF
 40501
 post spike = .04 ml of 8859
 in 10-1

Data Entry Comments:

Uploaded & Validated by:
Sam M. Pang
 01/06/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2235

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/06/99

13:33

Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	119700.
Stddev	1266.
%RSD	1.057

#1	119200.
#2	119400.
#3	119700.
#4	119900.
#5	117600.
#6	120100.
#7	118000.
#8	120500.
#9	121500.
#10	121300.

1-6-99

FILE # ALD106A.TRT

UL # 27475

588700 3012, 3013

.1-10-1025-10 40501 BT

64-105

SCANS : 983012
983013

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES ~~2236~~ TO ~~2240~~

Method: ACTINIDE Sample Name: cal blank

Operator: bjg

Comment: background

Run Time: 01/06/99 13:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3488	.3149	.3581	.3800	.3231
Stddev	.0474	.0342	.0336	.0191	.0314
%RSD	13.60	10.85	9.397	5.033	9.719

#1	.4243	.3505	.4012	.3551	.3412
#2	.2996	.2950	.3319	.3688	.3504
#3	.3403	.2667	.3863	.3909	.2713
#4	.3217	.3401	.3263	.4044	.3171
#5	.3583	.3220	.3447	.3809	.3356

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3700	.3920	.3975	.3679	.3736
Stddev	.0629	.0346	.0279	.0459	.0397
%RSD	17.00	8.830	7.013	12.49	10.62

#1	.4565	.3781	.4012	.3505	.3781
#2	.3642	.3642	.3504	.3550	.4241
#3	.2805	.3955	.4185	.3219	.3127
#4	.3631	.4504	.4182	.3677	.3769
#5	.3855	.3719	.3991	.4444	.3764

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4078	.3525
Stddev	.0847	.0685
%RSD	20.78	19.43

#1	.4519	.2767
#2	.4379	.4426
#3	.5013	.3909
#4	.2849	.2941
#5	.3628	.3583

Sample Name: cal. blank Run Time: 01/06/99 13:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	72.622
Stddev	.504
%RSD	.69349

#1	72.284
#2	72.308
#3	72.485
#4	72.529
#5	73.502

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.010 ppm cal

Run Time: 01/06/99 13:45 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2737	.2770	171.9	255.1	.6764
Stddev	.0211	.0290	6.4	12.5	.0523
%RSD	7.710	10.46	3.737	4.916	7.726

#1	.2480	.2595	164.5	237.8	.6639
#2	.2692	.2375	166.4	248.3	.5977
#3	.2842	.3011	178.9	270.6	.7380
#4	.3036	.2799	177.4	257.4	.6781
#5	.2637	.3070	172.3	261.1	.7045

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.9410	.4031	.2707	.3098	.3273
Stddev	.0377	.0386	.0467	.0244	.0433
%RSD	4.009	9.570	17.26	7.867	13.22

#1	.9119	.4655	.2175	.3434	.3129
#2	.9302	.3761	.2375	.2810	.2613
#3	.9797	.3902	.2842	.3181	.3732
#4	.9817	.3706	.2760	.2917	.3548
#5	.9013	.4132	.3385	.3148	.3345

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3045	.2732
Stddev	.0454	.0358
%RSD	14.91	13.12

#1	.3510	.3053
#2	.2454	.2613
#3	.3350	.2714
#4	.2681	.3075
#5	.3227	.2204

Sample Name: cal std Run Time: 01/06/99 13:45

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	83.882
Stddev	3.211
%RSD	3.8283

#1	87.359
#2	84.208
#3	78.596
#4	84.550
#5	84.697

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 01/06/99 13:50 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3222	.5005	1701.	2459.	4.129
Stddev	.0326	.0266	13.	51.	.136
%RSD	10.11	5.306	.7587	2.076	3.281

#1	.3276	.5368	1710.	2396.	4.144
#2	.3665	.5044	1693.	2494.	4.079
#3	.3158	.5108	1684.	2427.	4.191
#4	.3255	.4824	1702.	2457.	3.934
#5	.2753	.4680	1717.	2523.	4.298

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.482	.9686	.2623	.2656	.2624
Stddev	.195	.0459	.0270	.0280	.0295
%RSD	3.005	4.744	10.31	10.54	11.23

#1	6.485	.9473	.2329	.2960	.2881
#2	6.231	.9222	.2522	.2798	.2168
#3	6.546	1.041	.3002	.2534	.2885
#4	6.758	.9492	.2471	.2236	.2628
#5	6.391	.9832	.2792	.2753	.2556

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2553	.2617
Stddev	.0394	.0407
%RSD	15.42	15.55

#1	.2842	.2329
#2	.1931	.3232
#3	.2768	.2261
#4	.2393	.2824
#5	.2832	.2438

Sample Name: cal std Run Time: 01/06/99 13:50

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.854
Stddev	.409
%RSD	.48229

#1	84.452
#2	84.581
#3	85.492
#4	84.988
#5	84.758

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact. lcs

Run Time: 01/06/99 14:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2560	.2694	182.8	11.26	.6834
Stddev	.0267	.0198	2.2	.20	.0487
%RSD	10.43	7.354	1.209	1.803	7.120
#1	.2389	.2624	186.5	11.34	.6540
#2	.3003	.2457	181.7	11.34	.6787
#3	.2483	.2996	181.6	11.37	.7686
#4	.2591	.2749	181.1	10.90	.6557
#5	.2331	.2647	182.9	11.37	.6598

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	341.4	380.9	11.35	.3267	.2914
Stddev	6.7	3.3	.30	.0346	.0422
%RSD	1.960	.8767	2.667	10.60	14.50
#1	345.3	383.2	11.08	.3525	.3094
#2	330.9	379.2	11.02	.3042	.2535
#3	348.6	383.6	11.72	.3469	.2523
#4	341.5	375.8	11.36	.3534	.3534
#5	340.6	382.8	11.56	.2766	.2884

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	484.7	.2913
Stddev	6.3	.0283
%RSD	1.306	9.703
#1	492.0	.2781
#2	485.8	.3277
#3	489.2	.3153
#4	477.8	.2709
#5	478.6	.2647

Sample Name: icv Run Time: 01/06/99 14:02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	84.881
Stddev	.431
%RSD	.50773

#1	85.114
#2	85.457
#3	84.568
#4	84.894
#5	84.372

2244

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% nitric

Run Time: 01/06/99 14:11 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2336	.2563	.3064	.2941	.2347
Stddev	.0282	.0206	.0453	.0367	.0341
%RSD	12.08	8.026	14.79	12.48	14.54

#1	.2745	.2269	.2562	.3403	.2232
#2	.2097	.2796	.2874	.2447	.1864
#3	.2264	.2456	.2802	.2878	.2802
#4	.2081	.2686	.3480	.3178	.2383
#5	.2492	.2606	.3603	.2798	.2453

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2639	.3314	.2954	.2739	.2977
Stddev	.0361	.0482	.0597	.0228	.0169
%RSD	13.67	14.54	20.22	8.334	5.675

#1	.2306	.3220	.2159	.3001	.3074
#2	.2874	.3806	.3417	.2485	.2874
#3	.2188	.3761	.2495	.2687	.2917
#4	.2913	.3140	.3518	.2573	.2800
#5	.2913	.2645	.3181	.2951	.3220

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2994	.2590
Stddev	.0411	.0164
%RSD	13.72	6.341

#1	.2964	.2525
#2	.3573	.2330
#3	.2418	.2725
#4	.2951	.2686
#5	.3066	.2683

Sample Name: icb Run Time: 01/06/99 14:11

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	87.769
Stddev	2.022
%RSD	2.3033
#1	91.086
#2	85.835
#3	86.853
#4	88.111
#5	86.962

2246

Method: ACTINIDE Sample Name: s98t003012

Operator: bjg

Comment: 40501 df

Run Time: 01/06/99 14:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2332	.2591	.2968	.5781	.2632
Stddev	.0172	.0393	.0257	.0651	.0297
%RSD	7.380	15.16	8.660	11.26	11.29

#1	.2562	.2791	.2715	.5429	.2600
#2	.2169	.2287	.2997	.6941	.2445
#3	.2151	.2071	.2948	.5537	.2310
#4	.2391	.2803	.3380	.5441	.2721
#5	.2387	.3004	.2798	.5555	.3086

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2704	.3255	.2435	.2875	.2648
Stddev	.0206	.0594	.0108	.0181	.0239
%RSD	7.601	18.26	4.439	6.297	9.016

#1	.2600	.2600	.2447	.2829	.2715
#2	.2524	.2879	.2603	.2800	.2761
#3	.2629	.3426	.2430	.3187	.2948
#4	.3050	.3215	.2308	.2844	.2350
#5	.2716	.4156	.2387	.2716	.2469

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2811	.2848
Stddev	.0494	.0332
%RSD	17.57	11.67

#1	.3212	.2868
#2	.3234	.3392
#3	.2749	.2669
#4	.2844	.2514
#5	.2016	.2798

Sample Name: s98t003012 Run Time: 01/06/99 14:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	83.450
Stddev	2.637
%RSD	3.1595

#1	87.183
#2	84.520
#3	83.679
#4	80.864
#5	81.003

Method: ACTINIDE Sample Name: s98t003012_d

Operator: bjg

Comment: 40501 df

Run Time: 01/06/99 14:22 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2486	.2220	.2736	.5471	.2637
Stddev	.0397	.0172	.0295	.0258	.0518
%RSD	15.98	7.743	10.78	4.707	19.63

#1	.2180	.2386	.2632	.5388	.2715
#2	.2019	.2019	.3239	.5889	.2818
#3	.2625	.2333	.2458	.5291	.1916
#4	.2582	.2049	.2664	.5532	.2418
#5	.3025	.2311	.2689	.5253	.3320

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2596	.2943	.2543	.2679	.2594
Stddev	.0411	.0190	.0219	.0274	.0317
%RSD	15.84	6.445	8.616	10.23	12.21

#1	.2550	.3002	.2673	.2262	.2961
#2	.3239	.2944	.2187	.2986	.2902
#3	.2666	.2666	.2750	.2583	.2458
#4	.2131	.2910	.2500	.2746	.2377
#5	.2395	.3194	.2605	.2815	.2269

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2644	.2711
Stddev	.0386	.0457
%RSD	14.61	16.87

#1	.2262	.1974
#2	.3197	.2776
#3	.2541	.2750
#4	.2869	.3237
#5	.2353	.2815

2249

Sample Name: s98t003012_d Run Time: 01/06/99 14:22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	80.193
Stdddev	.965
%RSD	1.2028

#1	81.044
#2	79.247
#3	80.008
#4	81.341
#5	79.326

Method: ACTINIDE Sample Name: s98t003012_a

Operator: bjg

Comment: 40501 df

Run Time: 01/06/99 14:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2368	.2887	304.9	2.408	.9133
Stddev	.0179	.0475	3.6	.080	.0635
%RSD	7.553	16.46	1.175	3.333	6.953
#1	.2352	.2310	300.1	2.369	.8401
#2	.2076	.2796	305.5	2.487	.8812
#3	.2420	.3057	305.6	2.378	.9001
#4	.2435	.3589	309.9	2.311	1.008
#5	.2556	.2683	303.2	2.496	.9371

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	616.1	693.9	20.40	.3861	.2834
Stddev	5.7	19.2	.21	.0313	.0256
%RSD	.9210	2.764	1.037	8.100	9.049
#1	607.2	661.1	20.04	.3865	.2940
#2	622.4	707.0	20.46	.3898	.2923
#3	616.7	708.3	20.54	.3694	.2378
#4	619.0	696.7	20.43	.3503	.2991
#5	615.2	696.5	20.56	.4345	.2939

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	897.1	.2675
Stddev	11.5	.0451
%RSD	1.283	16.87
#1	884.8	.1974
#2	905.6	.2881
#3	911.5	.2675
#4	896.2	.3204
#5	887.3	.2641

2251

Sample Name: s98t003012_a Run Time: 01/06/99 14:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	78.565
Stddev	.505
%RSD	.64310

#1	79.352
#2	78.678
#3	78.510
#4	78.025
#5	78.259

Method: ACTINIDE Sample Name: s98t003013

Operator: bjg

Comment: 40501 df

Run Time: 01/06/99 14:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2351	.2567	.2779	.5732	.2613
Stddev	.0207	.0209	.0299	.0223	.0232
%RSD	8.787	8.142	10.76	3.894	8.880

#1	.2401	.2608	.2898	.5754	.2401
#2	.2229	.2542	.2816	.5788	.3011
#3	.2664	.2545	.2863	.5368	.2545
#4	.2115	.2278	.3051	.5980	.2563
#5	.2347	.2865	.2268	.5769	.2546

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2664	.4678	.2561	.2744	.2957
Stddev	.0187	.0588	.0243	.0271	.0499
%RSD	7.017	12.58	9.479	9.867	16.88

#1	.2732	.4512	.2608	.2856	.3311
#2	.2620	.5319	.2229	.2699	.2347
#3	.2942	.3817	.2903	.2704	.3539
#4	.2441	.5126	.2482	.2359	.2563
#5	.2586	.4615	.2586	.3103	.3024

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2867	.2512
Stddev	.0271	.0421
%RSD	9.443	16.78

#1	.3187	.2649
#2	.2816	.2503
#3	.3101	.3141
#4	.2563	.2197
#5	.2666	.2069

Sample Name: s98t003013 Run Time: 01/06/99 14:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	83.063
Stddev	1.837
%RSD	2.2116
#1	80.528
#2	85.230
#3	83.835
#4	81.939
#5	83.781

Method: ACTINIDE Sample Name: s98t003013_d

Operator: bjg

Comment: 40501 df

Run Time: 01/06/99 14:42 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2273	.2005	.2744	.5698	.2187
Stddev	.0188	.0403	.0258	.0479	.0319
%RSD	8.267	20.11	9.399	8.406	14.61

#1	.1996	.1774	.2624	.4989	.1885
#2	.2256	.1950	.2600	.6156	.2103
#3	.2354	.2240	.2468	.5619	.1898
#4	.2249	.1513	.2944	.6133	.2535
#5	.2512	.2550	.3083	.5594	.2512

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2606	.3752	.2599	.2591	.2456
Stddev	.0198	.0266	.0429	.0149	.0336
%RSD	7.603	7.096	16.52	5.757	13.70

#1	.2550	.3437	.2698	.2846	.1996
#2	.2753	.3709	.2370	.2562	.2447
#3	.2658	.3948	.1974	.2582	.2278
#4	.2290	.4089	.2944	.2494	.2780
#5	.2778	.3577	.3006	.2474	.2778

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2360	.2420
Stddev	.0319	.0552
%RSD	13.53	22.81

#1	.2587	.1848
#2	.2179	.1835
#3	.1936	.2544
#4	.2739	.2944
#5	.2359	.2930

2255

Sample Name: s98t003013_d Run Time: 01/06/99 14:42

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	86.857
Stddev	3.204
%RSD	3.6893

#1	90.191
#2	87.184
#3	87.797
#4	81.525
#5	87.590

2256

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact. lcs

Run Time: 01/06/99 14:50 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2247	.2467	169.6	10.37	.5801
Stddev	.0366	.0247	1.9	.44	.0545
%RSD	16.27	9.991	1.135	4.234	9.387

#1	.2418	.2758	166.5	9.869	.5856
#2	.1775	.2265	170.5	10.50	.6306
#3	.2728	.2349	170.1	10.95	.5040
#4	.2290	.2710	169.1	9.991	.5496
#5	.2026	.2255	171.6	10.52	.6306

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	323.2	362.5	10.84	.3075	.2384
Stddev	3.0	4.7	.22	.0478	.0261
%RSD	.9158	1.296	2.058	15.53	10.95

#1	322.5	359.5	10.50	.2796	.1965
#2	318.8	355.8	11.03	.2568	.2492
#3	325.4	365.2	10.72	.3827	.2652
#4	322.8	365.5	10.95	.3168	.2328
#5	326.3	366.7	11.00	.3019	.2484

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	473.5	.2567
Stddev	.2	.0330
%RSD	.0423	12.87

#1	473.7	.2720
#2	473.2	.2114
#3	473.5	.2387
#4	473.6	.2633
#5	473.3	.2981

Sample Name: ccv Run Time: 01/06/99 14:50

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	87.807
Stddev	.497
%RSD	.56598

#1	88.222
#2	88.282
#3	87.971
#4	87.344
#5	87.217

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% nitric

Run Time: 01/06/99 14:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2242	.2207	.2638	.2437	.2236
Stddev	.0332	.0331	.0121	.0314	.0374
%RSD	14.83	15.00	4.577	12.89	16.73

#1	.2178	.1993	.2695	.2141	.1883
#2	.1948	.1948	.2772	.2173	.2472
#3	.2310	.2461	.2613	.2802	.2348
#4	.1996	.2661	.2661	.2329	.1811
#5	.2777	.1973	.2448	.2741	.2668

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2363	.3402	.2249	.2363	.2489
Stddev	.0313	.0367	.0354	.0200	.0360
%RSD	13.26	10.80	15.72	8.456	14.45

#1	.1956	.2953	.2326	.2067	.3027
#2	.2210	.3858	.2659	.2435	.2247
#3	.2688	.3143	.1818	.2537	.2651
#4	.2292	.3401	.1959	.2255	.2403
#5	.2668	.3654	.2485	.2521	.2119

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2429	.2505
Stddev	.0338	.0307
%RSD	13.92	12.25

#1	.2289	.2030
#2	.2098	.2884
#3	.2423	.2575
#4	.2994	.2551
#5	.2339	.2485

2259

Sample Name: ccb Run Time: 01/06/99 14:57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	89.743
Stddev	1.243
%RSD	1.3851

#1	90.300
#2	88.991
#3	88.029
#4	90.176
#5	91.220

LABCORE Completed Worklist Report for Worklist# 27492

Analyst: kjt Instrument: ICPMS1

Book#: 48B57
49B57

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	MSPR-QC TC99	QC	0.02050	1.88e-02	91.707 % Recovery	
1	ICV	0	MSPR-QC SN117	QC	0.01925	2.22e-02	115.325 % Recovery	
1	ICV	0	MSPR-QC CS133	QC	0.02000	1.94e-02	97.000 % Recovery	
2	ICB	0	MSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2	ICB	0	MSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2	ICB	0	MSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2	ICB	0	MSPR-QC CS133	QC	1	<2.00e-6	ug/mL	
2	ICB	0	MSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2	ICB	0	MSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3	BLNK-PREP	0	MSPR-A1 TC99-A1	SOLID	1	<2.20e-5	ug/g	
3	BLNK-PREP	0	MSPR-A1 SN117-A1	SOLID	1	7.04e-03	0.007 ug/g	
3	BLNK-PREP	0	MSPR-A1 SN126-A1	SOLID	1	<1.65e-3	ug/g	
3	BLNK-PREP	0	MSPR-A1 CS133-A1	SOLID	1	<2.20e-5	ug/g	
3	BLNK-PREP	0	MSPR-A1 CS135-A1	SOLID	1	2.53e-05	2.530e-005 ug/g	
3	BLNK-PREP	0	MSPR-A1 CS137-A1	SOLID	1	4.18e-05	4.180e-005 ug/g	
4	SAMPLE	S98T002604	0 A	MSPR-A1 TC99-A1	SOLID	N/A	1.592e-01	7.41e-002 ug/g
4	SAMPLE	S98T002604	0 A	MSPR-A1 SN117-A1	SOLID	N/A <	5.555e+00	5.554 ug/g
4	SAMPLE	S98T002604	0 A	MSPR-A1 SN126-A1	SOLID	N/A <	5.555e+00	5.554 ug/g
4	SAMPLE	S98T002604	0 A	MSPR-A1 CS133-A1	SOLID	N/A	1.259e-01	7.41e-002 ug/g
4	SAMPLE	S98T002604	0 A	MSPR-A1 CS135-A1	SOLID	N/A <	7.406e-02	7.41e-002 ug/g
4	SAMPLE	S98T002604	0 A	MSPR-A1 CS137-A1	SOLID	N/A <	7.406e-02	7.41e-002 ug/g
5	DUP	S98T002604	0 A	MSPR-A1 TC99-A1	SOLID	1.59e-01	9.33e-02	52.081 RPD
5	DUP	S98T002604	0 A	MSPR-A1 SN117-A1	SOLID	<5.55e0	<5.38e0	RPD
5	DUP	S98T002604	0 A	MSPR-A1 SN126-A1	SOLID	<5.55e0	<5.38e0	RPD
5	DUP	S98T002604	0 A	MSPR-A1 CS133-A1	SOLID	1.26e-01	7.90e-02	45.854 RPD
5	DUP	S98T002604	0 A	MSPR-A1 CS135-A1	SOLID	<7.41e-2	<7.18e-2	RPD
5	DUP	S98T002604	0 A	MSPR-A1 CS137-A1	SOLID	<7.41e-2	<7.18e-2	RPD
6	SPK	S98T002604	0 A	MSPR-A1 TC99-A1	SOLID	0.02190	7.38e+02	3369863.014 % Recovery
6	SPK	S98T002604	0 A	MSPR-A1 SN117-A1	SOLID	0.01930	7.31e+02	3787564.767 % Recovery
6	SPK	S98T002604	0 A	MSPR-A1 CS133-A1	SOLID	0.02010	7.31e+02	3636815.920 % Recovery
7	SAMPLE	S98T002605	0 A	MSPR-A1 TC99-A1	SOLID	N/A	3.176e-01	7.14e-002 ug/g
7	SAMPLE	S98T002605	0 A	MSPR-A1 SN117-A1	SOLID	N/A <	5.353e+00	5.353 ug/g
7	SAMPLE	S98T002605	0 A	MSPR-A1 SN126-A1	SOLID	N/A <	5.353e+00	5.353 ug/g
7	SAMPLE	S98T002605	0 A	MSPR-A1 CS133-A1	SOLID	N/A	2.748e-01	7.14e-002 ug/g
7	SAMPLE	S98T002605	0 A	MSPR-A1 CS135-A1	SOLID	N/A	8.922e-02	7.14e-002 ug/g
7	SAMPLE	S98T002605	0 A	MSPR-A1 CS137-A1	SOLID	N/A	9.993e-02	7.14e-002 ug/g
8	DUP	S98T002605	0 A	MSPR-A1 TC99-A1	SOLID	3.18e-01	3.35e-01	5.207 RPD
8	DUP	S98T002605	0 A	MSPR-A1 SN117-A1	SOLID	<5.35e0	<5.52e0	RPD
8	DUP	S98T002605	0 A	MSPR-A1 SN126-A1	SOLID	<5.35e0	<5.52e0	RPD
8	DUP	S98T002605	0 A	MSPR-A1 CS133-A1	SOLID	2.75e-01	3.09e-01	11.644 RPD
8	DUP	S98T002605	0 A	MSPR-A1 CS135-A1	SOLID	8.92e-02	9.94e-02	10.817 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27492

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
8 DUP	S98T002605	0 A	QMSPR-A1 CS137-A1	SOLID	9.99e-02	1.22e-01	19.919	RPD
9 SAMPLE	S98T002606	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	5.464e-01	7.54e-002	ug/g
9 SAMPLE	S98T002606	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	6.594e+00	5.652	ug/g
9 SAMPLE	S98T002606	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	5.652e+00	5.652	ug/g
9 SAMPLE	S98T002606	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	5.389e-01	7.54e-002	ug/g
9 SAMPLE	S98T002606	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	3.429e-01	7.54e-002	ug/g
9 SAMPLE	S98T002606	0 A	QMSPR-A1 CS137-A1	SOLID	N/A	9.948e-01	7.54e-002	ug/g
10 DUP	S98T002606	0 A	QMSPR-A1 TC99-A1	SOLID	5.46e-01	6.16e-01	12.048	RPD
10 DUP	S98T002606	0 A	QMSPR-A1 SN117-A1	SOLID	6.59e+00	<5.57e0		RPD
10 DUP	S98T002606	0 A	QMSPR-A1 SN126-A1	SOLID	<5.65e0	<5.57e0		RPD
10 DUP	S98T002606	0 A	QMSPR-A1 CS133-A1	SOLID	5.39e-01	6.35e-01	16.354	RPD
10 DUP	S98T002606	0 A	QMSPR-A1 CS135-A1	SOLID	3.43e-01	3.86e-01	11.797	RPD
10 DUP	S98T002606	0 A	QMSPR-A1 CS137-A1	SOLID	9.95e-01	1.15e+00	14.452	RPD
11 CCV	0		QMSPR-QC TC99	QC	0.02050	1.07e-02	91.220	% Recovery
11 CCV	0		QMSPR-QC SN117	QC	0.01925	2.21e-02	114.805	% Recovery
11 CCV	0		QMSPR-QC CS133	QC	0.02000	1.92e-02	96.000	% Recovery
12 CCB	0		QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
12 CCB	0		QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
12 CCB	0		QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27492

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sally M. King 03/08/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27492

Analyst: KST Instrument: ICPMS1 _____ Book# 48857Method: LA-506-101 Rev/Mod A-049857

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSPR-QC	QC		
2	ICB		@MSPR-QC	QC		
3	BLNK-PREP		@MSPR-A1	SOLID		
4	SAMPLE	S98T002604 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
5	DUP	S98T002604 0 A	@MSPR-A1	SOLID		
6	SPK-POST SP 2/8/99	S98T002604 0 A	@MSPR-A1	SOLID		
7	SAMPLE	S98T002605 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
8	DUP	S98T002605 0 A	@MSPR-A1	SOLID		
9	SAMPLE	S98T002606 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
10	DUP	S98T002606 0 A	@MSPR-A1	SOLID		
11	CCV		@MSPR-QC	QC		
12	CCB		@MSPR-QC	QC		

Final page for worklist # 27492

Analyst Signature KS Phornthong Date 03-05-99

Analyst Signature _____ Date _____

Data Entry Comments:

Uploaded + Validated by:

Sail N. Pany03/08/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/05/99 09:58 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	172100.
Stddev	1302.
%RSD	.7568

#1	172500.
#2	173000.
#3	169800.
#4	172800.
#5	172100.
#6	169800.
#7	173800.
#8	172100.
#9	172200.
#10	173000.

KJ Thumdeke 03-05-99

filepr0305a.txt

Blank-Prep	1-10	11 dP
9a5002604	.056-10	201 dP
-d	.05-10	201 dP
2604-a		
-a		
2605		
2605-d		
2606		
2606-d		

Scan pr 2604
pr 2605
pr 2606SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2244 TO 2285Spikes used .1ml of msprvsk 44857
.25ml of 10 ppm S

HNF-1663 REV. 0

Method: PRV

Sample Name: Blank

Operator:

Comment:

Run Time: 03/05/99 10:00 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.413	50.30	1.287	9.202	1.739
Stddev	.061	3.75	.253	.642	.147
%RSD	4.337	7.459	19.63	6.977	8.484
#1	1.375	47.00	1.633	8.134	1.575
#2	1.514	45.50	1.428	9.368	1.742
#3	1.422	52.50	.9668	9.697	1.763
#4	1.358	53.50	1.213	9.130	1.647
#5	1.396	53.00	1.196	9.683	1.965

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.877	2.460	3.284	12.98
Stddev	.326	.048	.202	.56
%RSD	17.35	1.937	6.145	4.316
#1	2.349	2.492	3.322	13.12
#2	1.457	2.513	3.427	13.74
#3	1.735	2.389	3.185	12.28
#4	1.907	2.456	3.496	12.60
#5	1.937	2.449	2.990	13.16

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.482
Stddev	.109
%RSD	.62625

#1	17.457
#2	17.506
#3	17.583
#4	17.306
#5	17.556

Method: PRV

Sample Name: PRV CAL

Operator:

Comment:

Run Time: 03/05/99 10:03 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Cs[133]
Line	99/pulse	133/pulse
Units	Cts/S	Cts/S
Avg	23640.	20560.
Stddev	1438.	1183.
%RSD	6.083	5.752

#1	26150.	22650.
#2	23490.	20360.
#3	22860.	19980.
#4	22950.	19860.
#5	22710.	19950.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.297
Stddev	.807
%RSD	4.6677

#1	15.875
#2	17.440
#3	17.830
#4	17.653
#5	17.689

Method: PRV

Sample Name: PRV1 CAL

Operator:

Comment:

Run Time: 03/05/99 10:07 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Ru[102]	Te[125]	Cs[135]	Cs[137]
Line	102/pulse	125/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	669600.	1598.	7064.	12260.
Stddev	9553.	47.	252.	384.
%RSD	1.427	2.945	3.572	3.129

#1	686200.	1679.	7510.	12940.
#2	668600.	1573.	6939.	12180.
#3	662100.	1587.	6931.	12020.
#4	666200.	1560.	6921.	12020.
#5	664900.	1593.	7021.	12140.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	75300.
Stddev	2162.
%RSD	2.871

#1	79160.
#2	74350.
#3	74390.
#4	74090.
#5	74510.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.539
Stddev	.480
%RSD	2.7373

#1	16.687
#2	17.766
#3	17.808
#4	17.782
#5	17.650

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1199.6804	1.4128	1.0000000	03/05/99 10:04:16
Ru[102]	6695.4121	50.3000	1.0000000	03/05/99 10:09:17
Te[125]	15.9711	1.2874	1.0000000	03/05/99 10:09:17
Sn[117]	962.8264	9.2025	1.0000000	03/05/99 10:02:48
Sn[126]	963.2141	1.7385	1.0000000	03/05/99 10:02:48
Cs[133]	1027.7765	1.8767	1.0000000	03/05/99 10:04:16
Cs[135]	1069.9445	2.4598	1.0000000	03/05/99 10:09:17
Cs[137]	1094.3969	3.2842	1.0000000	03/05/99 10:09:17
Ba[138]	752.8637	12.9790	1.0000000	03/05/99 10:09:17

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV-SN CAL

Operator:

Comment:

Run Time: 03/05/99 10:12 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Sn[117]	Sn[126]
Line	117/pulse	126/pulse
Units	Cts/S	Cts/S
Avg	18350.	18350.
Stddev	387.	387.
%RSD	2.108	2.108

#1	18250.	18250.
#2	18700.	18700.
#3	18820.	18820.
#4	18000.	18000.
#5	17990.	17990.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.408
Stddev	.248
%RSD	1.4264

#1	17.492
#2	17.090
#3	17.206
#4	17.595
#5	17.658

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1199.6804	1.4128	1.0000000	03/05/99 10:04:16
Ru[102]	6695.4121	50.3000	1.0000000	03/05/99 10:09:17
Te[125]	15.9711	1.2874	1.0000000	03/05/99 10:09:17
Sn[117]	952.9421	9.2025	1.0000000	03/05/99 10:13:29
Sn[126]	953.3298	1.7385	1.0000000	03/05/99 10:13:29
Cs[133]	1027.7765	1.8767	1.0000000	03/05/99 10:04:16
Cs[135]	1069.9445	2.4598	1.0000000	03/05/99 10:09:17
Cs[137]	1094.3969	3.2842	1.0000000	03/05/99 10:09:17
Ba[138]	752.8637	12.9790	1.0000000	03/05/99 10:09:17

HNF-1663 REV. 0

Method: PRV Sample Name: icv

Operator: kjt

Comment: msprv lcs

Run Time: 03/05/99 10:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.77	.0048	-.0080	1.199	.0000
Stddev	.23	.0007	.0088	.124	.0002
%RSD	1.203	13.82	110.9	10.36	1048.

#1	19.11	.0052	.0016	1.029	-.0001
#2	18.81	.0050	-.0081	1.122	.0000
#3	18.57	.0044	-.0047	1.230	-.0001
#4	18.81	.0039	-.0224	1.271	-.0001
#5	18.56	.0056	-.0063	1.343	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.36	.0067	.0118	.1089
Stddev	.21	.0007	.0004	.0022
%RSD	1.080	10.51	3.433	2.050

#1	19.56	.0067	.0113	.1127
#2	19.20	.0073	.0122	.1077
#3	19.14	.0074	.0118	.1073
#4	19.30	.0063	.0121	.1077
#5	19.59	.0057	.0115	.1090

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.358
Stddev	.199
%RSD	1.0830

#1	18.281
#2	18.577
#3	18.548
#4	18.274
#5	18.111

Method: PRV   Sample Name: icv
 Comment: 19.25 ppb Sn
 Run Time: 03/05/99 10:23 Type: Unk

Operator: kjt
 Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0883	.0002	.0085	22.17	.0003
Stddev	.1037	.0015	.0263	.49	.0003
%RSD	117.5	843.2	309.4	2.225	109.1

#1	.2671	.0009	.0526	23.04	.0006
#2	.0894	-.0021	.0017	21.87	-.0002
#3	.0409	.0020	-.0168	22.09	.0001
#4	.0265	.0000	-.0027	21.89	.0005
#5	.0176	.0000	.0077	21.97	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0074	.0266	.0446	.4218
Stddev	.0014	.0006	.0019	.0105
%RSD	19.07	2.161	4.355	2.491

#1	.0097	.0274	.0464	.4403
#2	.0064	.0259	.0445	.4149
#3	.0063	.0261	.0453	.4181
#4	.0070	.0265	.0456	.4200
#5	.0074	.0269	.0414	.4159

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	16.066
Stddev	.233
%RSD	1.4522

#1	15.745
#2	16.359
#3	16.195
#4	16.073
#5	15.957

Method: PRV

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/05/99 10:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0017	.0128	-.0046	.0207	-.0001
Stddev	.0004	.0024	.0108	.0009	.0002
%RSD	21.18	18.67	233.3	4.540	252.4
#1	.0020	.0168	-.0180	.0217	-.0001
#2	.0019	.0129	.0028	.0214	.0002
#3	.0020	.0124	-.0100	.0203	-.0003
#4	.0013	.0114	.0094	.0208	-.0001
#5	.0013	.0106	-.0074	.0193	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0003	-.0001	-.0002	.0006
Stddev	.0003	.0003	.0004	.0005
%RSD	110.3	412.7	245.3	76.35
#1	.0008	-.0003	-.0006	.0009
#2	.0001	-.0004	-.0002	.0012
#3	.0001	-.0002	-.0003	.0003
#4	.0001	.0004	.0003	.0000
#5	.0004	.0002	.0001	.0008

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.541
Stddev	.151
%RSD	.81193

#1	18.515
#2	18.762
#3	18.612
#4	18.437
#5	18.382

HNF-1663 REV. 0

Method: PRV

Sample Name: blnk-prep

Operator: kjt

Comment: 11 df

Run Time: 03/05/99 10:34 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0012	.0062	-.0059	.6397	.0000
Stddev	.0002	.0013	.0065	.0207	.000
%RSD	13.12	21.33	109.3	3.239	824.3

#1	.0011	.0062	-.0001	.6090	.0003
#2	.0011	.0077	-.0003	.6373	-.0004
#3	.0014	.0056	-.0095	.6381	.0004
#4	.0012	.0070	-.0152	.6661	-.0005
#5	.0010	.0043	-.0045	.6482	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0003	.0023	.0038	.0348
Stddev	.0002	.0002	.0003	.0024
%RSD	53.95	6.611	8.048	6.900

#1	.0004	.0021	.0038	.0335
#2	.0001	.0023	.0043	.0352
#3	.0004	.0021	.0038	.0386
#4	.0003	.0025	.0038	.0345
#5	.0006	.0024	.0034	.0322

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.354
Stddev	.142
%RSD	.77305

#1	18.270
#2	18.320
#3	18.497
#4	18.182
#5	18.503

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002604

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0043	.0101	-.0062	.0370	-.0001
Stddev	.0006	.0009	.0120	.0018	.0001
%RSD	13.52	8.533	193.8	4.770	136.8

#1	.0039	.0100	-.0005	.0390	-.0002
#2	.0035	.0100	-.0189	.0379	.0001
#3	.0050	.0112	.0063	.0344	-.0003
#4	.0045	.0103	.0014	.0361	-.0001
#5	.0045	.0088	-.0193	.0374	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0034	.0008	.0017	-.0010
Stddev	.0004	.0001	.0002	.0004
%RSD	11.85	17.95	14.23	45.06

#1	.0035	.0008	.0017	-.0008
#2	.0038	.0009	.0019	-.0009
#3	.0029	.0009	.0017	-.0005
#4	.0030	.0008	.0020	-.0011
#5	.0036	.0006	.0014	-.0017

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.850
Stddev	.147
%RSD	.78082

#1	18.769
#2	18.782
#3	18.725
#4	19.093
#5	18.880

Method: PRV

Sample Name: s98t002604_d

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0026	.0064	-.0033	.0579	.0000
Stddev	.0003	.0013	.0105	.0011	.000
%RSD	11.27	20.79	318.3	1.882	2709.

#1	.0024	.0085	.0026	.0561	.0002
#2	.0028	.0062	.0073	.0589	.0000
#3	.0028	.0058	.0027	.0578	.0002
#4	.0029	.0050	-.0127	.0586	-.0001
#5	.0023	.0065	-.0164	.0579	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0022	.0005	.0006	-.0028
Stddev	.0003	.0002	.0002	.0013
%RSD	15.36	41.43	42.66	46.92

#1	.0023	.0006	.0005	-.0035
#2	.0027	.0005	.0009	-.0015
#3	.0020	.0008	.0007	-.0040
#4	.0018	.0002	.0003	-.0012
#5	.0023	.0003	.0004	-.0038

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.878
Stddev	.083
%RSD	.43985

#1	18.813
#2	18.874
#3	18.797
#4	18.902
#5	19.005

Method: PRV

Sample Name: s98t002604_a

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:45 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.72	.0076	-.0035	3.707	-.0002
Stddev	.21	.0019	.0155	.039	.0001
%RSD	1.043	25.37	444.1	1.061	76.63

#1	19.94	.0099	-.0204	3.761	-.0003
#2	19.46	.0067	-.0001	3.717	.0000
#3	19.63	.0068	-.0025	3.674	-.0002
#4	19.65	.0053	-.0144	3.718	-.0002
#5	19.92	.0094	.0199	3.662	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.54	.0021	.0035	.0121
Stddev	.23	.0004	.0003	.0002
%RSD	1.188	19.87	7.863	1.284

#1	19.72	.0014	.0035	.0119
#2	19.29	.0023	.0031	.0121
#3	19.83	.0025	.0039	.0124
#4	19.51	.0022	.0034	.0122
#5	19.36	.0019	.0037	.0121

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.903
Stddev	.135
%RSD	.71173

#1	18.717
#2	19.050
#3	18.830
#4	18.911
#5	19.007

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002604_a

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0112	.0041	.0034	19.52	-.0001
Stddev	.0006	.0008	.0177	.46	.0003
%RSD	5.508	18.76	526.6	2.342	374.5
#1	.0119	.0031	-.0030	20.23	.0000
#2	.0116	.0039	.0328	19.44	.0002
#3	.0114	.0038	-.0078	19.09	-.0006
#4	.0106	.0050	.0060	19.18	.0001
#5	.0105	.0047	-.0112	19.68	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0058	.0028	.0025	.0023
Stddev	.0004	.0003	.0002	.0012
%RSD	7.344	9.911	6.831	54.71
#1	.0065	.0026	.0028	.0040
#2	.0058	.0030	.0024	.0023
#3	.0056	.0032	.0023	.0007
#4	.0053	.0025	.0025	.0017
#5	.0060	.0027	.0025	.0026

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.243
Stddev	.331
%RSD	1.8131

#1	17.758
#2	18.500
#3	18.487
#4	18.429
#5	18.039

Method: PRV

Sample Name: s98t002605

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0089	.0098	-.0038	.0318	.0001
Stddev	.0007	.0009	.0101	.0016	.0003
%RSD	8.405	9.388	263.7	5.116	278.8

#1	.0096	.0085	.0113	.0340	.0005
#2	.0090	.0110	-.0090	.0303	.0002
#3	.0093	.0102	-.0067	.0304	-.0001
#4	.0089	.0099	-.0149	.0316	.0002
#5	.0077	.0094	.0002	.0329	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0077	.0025	.0028	-.0015
Stddev	.0008	.0003	.0007	.0013
%RSD	10.85	13.28	23.47	87.05

#1	.0083	.0028	.0038	.0000
#2	.0085	.0023	.0022	-.0014
#3	.0079	.0024	.0031	-.0037
#4	.0075	.0029	.0030	-.0010
#5	.0064	.0022	.0022	-.0016

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.722
Stddev	.579
%RSD	3.0929

#1	17.706
#2	18.788
#3	19.065
#4	19.071
#5	18.979

HNF-1663 REV.0

Method: PRV

Sample Name: s98t002605_d

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 10:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0091	.0093	-.0124	.0345	.0001
Stddev	.0003	.0008	.0059	.0005	.0003
%RSD	3.179	8.192	47.69	1.409	345.2
#1	.0086	.0091	-.0040	.0343	-.0002
#2	.0093	.0097	-.0096	.0344	.0005
#3	.0091	.0104	-.0156	.0352	-.0003
#4	.0093	.0083	-.0132	.0346	.0003
#5	.0093	.0092	-.0194	.0339	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0084	.0027	.0033	-.0032
Stddev	.0005	.0002	.0002	.0009
%RSD	5.735	8.813	4.680	27.58
#1	.0090	.0024	.0033	-.0022
#2	.0079	.0028	.0030	-.0039
#3	.0087	.0024	.0031	-.0024
#4	.0085	.0027	.0034	-.0031
#5	.0079	.0029	.0034	-.0042

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.906
Stddev	.110
%RSD	.58317

#1	18.790
#2	18.971
#3	18.795
#4	19.039
#5	18.937

Method: PRV

Sample Name: s98t002606

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 11:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0145	.0551	-.0028	.1750	.0006
Stddev	.0005	.0022	.0094	.0027	.0003
%RSD	3.704	4.077	332.3	1.570	56.69
#1	.0140	.0569	-.0018	.1751	.0004
#2	.0146	.0533	-.0066	.1757	.0001
#3	.0138	.0525	.0111	.1757	.0008
#4	.0148	.0578	-.0020	.1706	.0008
#5	.0151	.0551	-.0149	.1781	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0143	.0091	.0264	.0698
Stddev	.0005	.0007	.0010	.0015
%RSD	3.829	8.002	3.622	2.217
#1	.0149	.0090	.0259	.0706
#2	.0146	.0103	.0260	.0695
#3	.0138	.0084	.0268	.0720
#4	.0136	.0088	.0278	.0687
#5	.0145	.0088	.0254	.0682

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.077
Stddev	.037
%RSD	.19546

#1	19.079
#2	19.034
#3	19.117
#4	19.109
#5	19.044

Analysis Report

03/05/99 11:08:35

page 1

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002606_d

Operator: kjt

Comment: 201 df

Run Time: 03/05/99 11:06 Type: Unk

Mode: CONC

Corr.Fact: 1.000000.

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0166	.0609	-.0043	.0458	.0005
Stddev	.0009	.0022	.0182	.0020	.0002
%RSD	5.491	3.542	420.3	4.364	46.01
#1	.0162	.0603	.0186	.0461	.0004
#2	.0171	.0610	-.0146	.0478	.0004
#3	.0180	.0644	-.0259	.0430	.0006
#4	.0157	.0586	-.0097	.0446	.0002
#5	.0161	.0601	.0099	.0474	.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0171	.0104	.0310	.0785
Stddev	.0004	.0006	.0020	.0014
%RSD	2.349	5.725	6.486	1.829
#1	.0176	.0106	.0318	.0806
#2	.0165	.0099	.0289	.0773
#3	.0174	.0106	.0310	.0773
#4	.0170	.0096	.0293	.0779
#5	.0172	.0111	.0336	.0792

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.827
Stddev	.172
%RSD	.91309

#1	18.617
#2	18.980
#3	18.889
#4	18.978
#5	18.673

Analysis Report

HNF-1663 REV. 0

Method: PRV Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/05/99 11:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.71	.0006	-.0100	1.158	-.0002
Stddev	.32	.0012	.0130	.133	.0003
%RSD	1.684	183.5	130.2	11.49	109.5

#1	19.26	-.0010	-.0097	.9731	-.0003
#2	18.53	.0022	-.0233	1.095	-.0005
#3	18.57	.0004	.0072	1.158	.0000
#4	18.51	.0006	-.0022	1.252	-.0004
#5	18.66	.0009	-.0219	1.312	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.15	.0066	.0109	.1087
Stddev	.44	.0006	.0005	.0043
%RSD	2.307	9.311	4.816	3.987

#1	19.89	.0075	.0117	.1151
#2	19.05	.0064	.0110	.1070
#3	18.74	.0060	.0103	.1110
#4	18.92	.0068	.0108	.1054
#5	19.13	.0062	.0108	.1048

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.602
Stddev	.302
%RSD	1.6255

#1	18.102
#2	18.586
#3	18.892
#4	18.770
#5	18.660

Analysis Report

HNF-1663 REV.0

Method: PRV Sample Name: ccv

Operator: kjt

Comment: 19.25 ppb Sn

Run Time: 03/05/99 11:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.4156	-.0007	-.0009	22.13	.0002
Stddev	.5407	.0007	.0157	.23	.0002
%RSD	130.1	93.17	1781.	1.041	137.7

#1	1.351	-.0002	-.0058	22.31	-.0001
#2	.4137	-.0004	-.0118	22.35	.0003
#3	.1600	-.0018	-.0172	21.79	-.0001
#4	.0931	-.0006	.0099	22.18	.0004
#5	.0607	-.0004	.0205	22.02	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0089	.0277	.0444	.4228
Stddev	.0016	.0010	.0016	.0077
%RSD	18.10	3.543	3.613	1.821

#1	.0115	.0268	.0462	.4321
#2	.0095	.0266	.0434	.4270
#3	.0081	.0278	.0427	.4116
#4	.0081	.0289	.0461	.4202
#5	.0075	.0283	.0438	.4230

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.943
Stddev	.297
%RSD	1.8627

#1	15.492
#2	15.921
#3	16.294
#4	15.909
#5	16.098

Method: PRV

Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/05/99 11:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0013	.0034	-.0102	.0178	.0000
Stddev	.0003	.0008	.0125	.0012	.000
%RSD	23.20	22.85	122.2	6.926	1730.
#1	.0017	.0035	-.0063	.0197	.0003
#2	.0013	.0025	-.0245	.0181	-.0001
#3	.0010	.0033	.0083	.0174	-.0005
#4	.0010	.0030	-.0105	.0177	.0003
#5	.0014	.0046	-.0181	.0162	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0005	.0000	.0000	.0017
Stddev	.0002	.000	.0004	.0008
%RSD	37.76	338.8	1706.	45.69
#1	.0006	.0001	-.0003	.0019
#2	.0006	-.0002	.0002	.0017
#3	.0002	.0001	.0001	.0004
#4	.0006	.0000	-.0004	.0021
#5	.0007	-.0002	.0005	.0023

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.961
Stddev	.114
%RSD	.63611

#1	18.113
#2	17.843
#3	17.959
#4	17.860
#5	18.031

LABCORE Completed Worklist Report for Worklist# 27493

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-5

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSU

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV		0	EMS-U-QC U235	QC	0.000142	1.43e-04	100.704 % Recovery	
1 ICV		0	EMS-U-QC U238	QC	0.020	2.07e-02	103.500 % Recovery	
2 ICB		0	EMS-U-QC U233	QC	1	<1.20e-5		ug/mL
2 ICB		0	EMS-U-QC U234	QC	1	<1.20e-5		ug/mL
2 ICB		0	EMS-U-QC U235	QC	1	<1.20e-5		ug/mL
2 ICB		0	EMS-U-QC U236	QC	1	<1.60e-5		ug/mL
2 ICB		0	EMS-U-QC U238	QC	1	<1.20e-5		ug/mL
3 STD-PREP		0	EMS-U-A1 U238-A1	SOLID	10.0	9.90e+00	99.000 % Recovery	
4 BLNK-PREP		0	EMS-U-A1 U233-A1	SOLID	1	<1.20e-4		ug/g
4 BLNK-PREP		0	EMS-U-A1 U234-A1	SOLID	1	<1.20e-4		ug/g
4 BLNK-PREP		0	EMS-U-A1 U235-A1	SOLID	1	<1.20e-4		ug/g
4 BLNK-PREP		0	EMS-U-A1 U236-A1	SOLID	1	<1.60e-4		ug/g
4 BLNK-PREP		0	EMS-U-A1 U238-A1	SOLID	1	<1.20e-4		ug/g
5 SAMPLE	S98T002604	0 A	EMS-U-A1 U233-A1	SOLID	N/A	< 4.444e-01	0.444	ug/g
5 SAMPLE	S98T002604	0 A	EMS-U-A1 U234-A1	SOLID	N/A	< 4.444e-01	0.444	ug/g
5 SAMPLE	S98T002604	0 A	EMS-U-A1 U235-A1	SOLID	N/A	< 4.444e-01	0.444	ug/g
5 SAMPLE	S98T002604	0 A	EMS-U-A1 U236-A1	SOLID	N/A	< 5.925e-01	0.593	ug/g
5 SAMPLE	S98T002604	0 A	EMS-U-A1 U238-A1	SOLID	N/A	< 7.110e-01	0.444	ug/g
6 DUP	S98T002604	0 A	EMS-U-A1 U233-A1	SOLID	<4.44e-1	<4.31e-1		RPD
6 DUP	S98T002604	0 A	EMS-U-A1 U234-A1	SOLID	<4.44e-1	<4.31e-1		RPD
6 DUP	S98T002604	0 A	EMS-U-A1 U235-A1	SOLID	<4.44e-1	<4.31e-1		RPD
6 DUP	S98T002604	0 A	EMS-U-A1 U236-A1	SOLID	<5.93e-1	<5.74e-1		RPD
6 DUP	S98T002604	0 A	EMS-U-A1 U238-A1	SOLID	7.11e-01	1.06e+00	39.413	RPD
7 SPK	S98T002604	0 A	EMS-U-A1 U235-A1	SOLID	4.0	3.93e+00	98.250 % Recovery	
7 SPK	S98T002604	0 A	EMS-U-A1 U238-A1	SOLID	2000.0	1.92e+03	96.000 % Recovery	
8 SAMPLE	S98T002605	0 A	EMS-U-A1 U233-A1	SOLID	N/A	< 4.283e-01	0.428	ug/g
8 SAMPLE	S98T002605	0 A	EMS-U-A1 U234-A1	SOLID	N/A	< 4.283e-01	0.428	ug/g
8 SAMPLE	S98T002605	0 A	EMS-U-A1 U235-A1	SOLID	N/A	< 4.283e-01	0.428	ug/g
8 SAMPLE	S98T002605	0 A	EMS-U-A1 U236-A1	SOLID	N/A	< 5.710e-01	0.571	ug/g
8 SAMPLE	S98T002605	0 A	EMS-U-A1 U238-A1	SOLID	N/A	< 3.626e+00	0.428	ug/g
9 DUP	S98T002605	0 A	EMS-U-A1 U233-A1	SOLID	<4.28e-1	<4.42e-1		RPD
9 DUP	S98T002605	0 A	EMS-U-A1 U234-A1	SOLID	<4.28e-1	<4.42e-1		RPD
9 DUP	S98T002605	0 A	EMS-U-A1 U235-A1	SOLID	<4.28e-1	<4.42e-1		RPD
9 DUP	S98T002605	0 A	EMS-U-A1 U236-A1	SOLID	<5.71e-1	<5.89e-1		RPD
9 DUP	S98T002605	0 A	EMS-U-A1 U238-A1	SOLID	3.63e+00	3.89e+00	6.915	RPD
10 SAMPLE	S98T002606	0 A	EMS-U-A1 U233-A1	SOLID	N/A	< 4.522e-01	0.452	ug/g
10 SAMPLE	S98T002606	0 A	EMS-U-A1 U234-A1	SOLID	N/A	< 4.522e-01	0.452	ug/g
10 SAMPLE	S98T002606	0 A	EMS-U-A1 U235-A1	SOLID	N/A	< 1.813e+00	0.452	ug/g
10 SAMPLE	S98T002606	0 A	EMS-U-A1 U236-A1	SOLID	N/A	< 6.029e-01	0.603	ug/g
10 SAMPLE	S98T002606	0 A	EMS-U-A1 U238-A1	SOLID	N/A	< 2.723e+02	0.452	ug/g
11 DUP	S98T002606	0 A	EMS-U-A1 U233-A1	SOLID	<4.52e-1	<4.45e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

HNF-1663 REV.0

LABCORE Completed Worklist Report for Worklist# 27493

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	DUP	S98T002606	0	A	QMSU-A1	U234-A1	SOLID	<4.52e-1	<4.45e-1	RPD
11	DUP	S98T002606	0	A	QMSU-A1	U235-A1	SOLID	1.81e+00	2.07e+00	13.402 RPD
11	DUP	S98T002606	0	A	QMSU-A1	U236-A1	SOLID	<6.03e-1	<5.94e-1	RPD
11	DUP	S98T002606	0	A	QMSU-A1	U238-A1	SOLID	2.72e+02	3.09e+02	12.737 RPD
12	CCV		0		QMSU-QC	U235	QC	0.000142	1.48e-04	104.225 % Recovery
12	CCV		0		QMSU-QC	U238	QC	0.020	2.12e-02	106.000 % Recovery
13	CCB		0		QMSU-QC	U233	QC	1	<1.20e-5	ug/mL
13	CCB		0		QMSU-QC	U234	QC	1	<1.20e-5	ug/mL
13	CCB		0		QMSU-QC	U235	QC	1	<1.20e-5	ug/mL
13	CCB		0		QMSU-QC	U236	QC	1	<1.60e-5	ug/mL
13	CCB		0		QMSU-QC	U238	QC	1	<1.20e-5	ug/mL

Final page for worklist# 27493

Analyst Signature

Date

Analyst Signature

Date

Sam H. Pang 3/01/99
_____
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27493

Analyst: BJG Instrument: ICPMS1 _____ Book# 990226-5

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 STD-PREP			@MSU-A1	SOLID		
4 BLNK-PREP			@MSU-A1	SOLID		
5 SAMPLE	S98T002604 0 A		@MSU-A1	SOLID	98000436 BY-105 (2)	
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1						
6 DUP	S98T002604 0 A		@MSU-A1	SOLID		
- SPK - post	2604					
7 SPK - pre	S98T002604 0 A		@MSU-A1	SOLID		
8 SAMPLE	S98T002605 0 A		@MSU-A1	SOLID	98000436 BY-105 (2)	
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1						
9 DUP	S98T002605 0 A		@MSU-A1	SOLID		
10 SAMPLE	S98T002606 0 A		@MSU-A1	SOLID	98000436 BY-105 (2)	
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1						
11 DUP	S98T002606 0 A		@MSU-A1	SOLID		
12 CCV			@MSU-QC	QC		
13 CCB			@MSU-QC	QC		

Final page for worklist # 27493

BJG 2-26-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

SCANS = 982604, 05, 06d

Uploaded + Validated by:
Sa M. Pang

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

03/01/99

2288

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 02/26/99 13:28 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/s
Avg	165500.
Stddev	1064.
%RSD	.6427

#1	164100.
#2	164800.
#3	165000.
#4	164400.
#5	167600.
#6	166900.
#7	165600.
#8	165600.
#9	165800.
#10	165600.

bjg
2-26-99
By-105 ISO-U
105-10-1
ACID DIGEST
FILE # 00226A.TXT
W.C.# 27483
SCANS = 982604
982605
982606d

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2287 TO 230.

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/26/99 13:30 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.617	6.903	6.726	7.131	7.192
Stddev	.814	.544	.298	.989	.708
%RSD	12.30	7.884	4.435	13.87	9.850

#1	6.180	7.762	7.020	7.070	7.663
#2	7.251	7.056	6.959	6.861	7.445
#3	7.155	6.624	6.479	8.751	6.140
#4	7.115	6.341	6.825	6.050	6.825
#5	5.385	6.731	6.346	6.923	7.885

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.292
Stddev	.109
%RSD	1.0607

#1	10.114
#2	10.275
#3	10.342
#4	10.330
#5	10.400

Method : ISOU

Standardzn Report

02/26/99 13:31:42

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	437.2802	6.6172	1.0000000	02/26/99 13:31:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.6172
100ppb U	0.711	0.711	0	0	317.52

U [233]

0
(S) IR³⁵⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	436.8787	6.9026	1.0000000	02/26/99 13:31:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.9026
100ppb U	0.711	0.711	0	0	317.52

U [234]

0
(S) IR³⁵⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	437.1275	6.7257	1.0000000	02/26/99 13:31:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.7257
100ppb U	0.711	0.711	0	0	317.52

U [235]

0
(S) IR³⁵⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	436.5574	7.1311	1.0000000	02/26/99 13:31:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	7.1311
100ppb U	0.711	0.711	0	0	317.52

U [236]

0
(S) IR³⁵⁰⁰¹⁰

Concentration

Method : ISOU

Standardzn Report

02/26/99 13:31:42

page 2

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [238]	398.5709	7.1916	1.0000000	02/26/99 13:31:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	7.1916
100ppb U	99.3	99.3	0	0	39585

U [238]

0

43550MITU
(s)IR

Concentration

2292

Analysis Report

02/26/99 13:33:50

page 1

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/26/99 13:32 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	331.0	331.0	331.0	331.0	41100.
Stddev	6.0	6.0	6.0	6.0	834.
%RSD	1.808	1.808	1.808	1.808	2.029
#1	334.1	334.1	334.1	334.1	40540.
#2	323.6	323.6	323.6	323.6	40870.
#3	329.1	329.1	329.1	329.1	40980.
#4	328.8	328.8	328.8	328.8	40560.
#5	339.4	339.4	339.4	339.4	42550.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.371
Stddev	.057
%RSD	.54923

#1	10.278
#2	10.384
#3	10.406
#4	10.424
#5	10.361

2293

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	456.2554	6.6172	1.0000000	02/26/99 13:33:48

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	6.6172
100ppb U	0.711	0.711	0	0	331.01

U [233]

0

(S) IR³⁷⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	455.8539	6.9026	1.0000000	02/26/99 13:33:48

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	6.9026
100ppb U	0.711	0.711	0	0	331.01

U [234]

0

(S) IR³⁷⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	456.1027	6.7257	1.0000000	02/26/99 13:33:48

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	6.7257
100ppb U	0.711	0.711	0	0	331.01

U [235]

0

(S) IR³⁷⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	455.5326	7.1311	1.0000000	02/26/99 13:33:48

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	7.1311
100ppb U	0.711	0.711	0	0	331.01

U [236]

0

(S) IR³⁷⁰⁰10

Concentration

2294

Method : ISOU

Standardzn Report

02/26/99 13:33:54

page 2

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	413.8259	7.1916	1.0000000	02/26/99 13:33:48

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	7.1916
100ppb U	99.3	99.3	0	0	41100

U [238]

0

45210MTO
(S)IR

Concentration

2295

Analysis Report

02/26/99 13:39:00

page 1

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 02/26/99 13:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0008	-.0015	-.0006	-.0010	.0001
Stddev	.0006	.0009	.0012	.0008	.0009
%RSD	86.01	63.76	217.7	77.96	603.0
#1	.0012	-.0027	-.0020	-.0020	-.0005
#2	.0006	-.0001	-.0002	-.0009	.0000
#3	-.0003	-.0015	-.0016	-.0018	.0010
#4	.0012	-.0011	.0010	-.0002	.0010
#5	.0011	-.0019	.0000	-.0003	-.0008

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.758
Stddev	.112
%RSD	1.0377

#1	10.626
#2	10.820
#3	10.659
#4	10.891
#5	10.791

2296

Analysis Report

02/26/99 13:41:53

page 1

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso-u lcs

Run Time: 02/26/99 13:40 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	.0010	.1433	.0021	20.72
Stddev	.0014	.0010	.0019	.0011	.16
%RSD	240.5	102.5	1.308	50.33	.7914
#1	.0027	.0018	.1405	.0019	20.58
#2	-.0006	.0022	.1425	.0006	20.81
#3	.0009	.0011	.1435	.0025	20.68
#4	.0007	-.0004	.1455	.0036	20.58
#5	-.0007	.0004	.1443	.0021	20.96

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.751
Stddev	.107
%RSD	.99882

#1	10.624
#2	10.670
#3	10.877
#4	10.837
#5	10.746

2297

Analysis Report

02/26/99 13:53:05

page 1

Method: ISOU Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 13:51 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0008	-.0015	-.0010	-.0005	.0001
Stddev	.0009	.0014	.0014	.0011	.0014
%RSD	113.6	89.23	138.5	219.0	1685.

#1	-.0020	-.0002	.0006	.0001	-.0010
#2	-.0009	-.0021	-.0019	-.0014	.0006
#3	.0004	-.0036	-.0015	-.0016	-.0017
#4	-.0003	-.0007	-.0025	.0010	.0013
#5	-.0014	-.0011	.0003	-.0006	.0013

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.901
Stddev	.107
%RSD	.97840

#1	10.765
#2	10.845
#3	10.880
#4	11.019
#5	10.998

2298

Analysis Report

02/26/99 13:58:14

page 1

Method: ISOU

Sample Name: prepstd

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 13:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0000	.1008	.0071	49.24
Stddev	.0014	.001	.0046	.0012	.82
%RSD	405.7	7051.	4.536	17.12	1.666
#1	-.0008	-.0011	.0966	.0072	48.25
#2	.0013	.0013	.1019	.0053	48.73
#3	-.0015	-.0010	.0954	.0066	49.73
#4	.0012	.0012	.1059	.0082	50.34
#5	.0016	-.0005	.1039	.0082	49.18

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.895
Stddev	.073
%RSD	.67327

#1	10.935
#2	10.911
#3	10.888
#4	10.775
#5	10.967

2299

Analysis Report

02/26/99 14:06:56

page 1

Method: ISOU Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 02/26/99 14:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	-.0004	-.0003	-.0007	.0014
Stddev	.0006	.0022	.0016	.0010	.0009
%RSD	357.9	555.1	604.7	150.9	67.11
#1	-.0003	-.0015	-.0004	-.0005	.0011
#2	-.0004	-.0004	.0005	.0004	.0000
#3	.0009	-.0022	-.0017	-.0001	.0017
#4	-.0009	.0033	.0016	-.0023	.0025
#5	-.0002	-.0012	-.0021	-.0010	.0015

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.002
Stddev	.061
%RSD	.55200

#1	11.023
#2	10.990
#3	10.912
#4	11.005
#5	11.080

2300

Analysis Report

02/26/99 14:18:52

page 1

Method: ISOU

Sample Name: S98t002604

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:17

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	-.0006	.0001	.0002	.0192
Stddev	.0018	.0011	.0008	.0014	.0019
%RSD	193.8	178.2	868.3	609.2	9.690

#1	.0004	-.0002	-.0002	.0027	.0193
#2	.0007	-.0010	.0015	-.0004	.0162
#3	.0037	.0011	.0002	.0000	.0202
#4	.0012	-.0017	-.0006	-.0011	.0212
#5	-.0014	-.0013	-.0004	.0000	.0190

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.9669
Stddev	.2125
%RSD	2.1325

#1	9.8215
#2	10.287
#3	9.7338
#4	9.9723
#5	10.020

2301

Analysis Report

02/26/99 14:21:23

page 1

Method: ISOU   Sample Name: s98t002604_d   Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 14:20 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	-.0009	.0001	-.0009	.0294
Stddev	.0010	.0016	.0010	.0014	.0024
%RSD	545.8	181.4	844.8	153.6	8.055

#1	-.0001	.0019	.0009	-.0009	.0301
#2	.0017	-.0020	-.0004	-.0020	.0275
#3	-.0010	-.0017	-.0014	-.0010	.0275
#4	.0005	-.0014	.0007	-.0020	.0289
#5	-.0002	-.0011	.0008	.0014	.0332

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.195
Stddev	.105
%RSD	1.0297

#1	10.106
#2	10.068
#3	10.233
#4	10.320
#5	10.247

2302

Analysis Report

02/26/99 14:24:58

page 1

Method: ISOU

Sample Name: s98t002604_a

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0005	.0001	.1379	.0019	19.69
Stddev	.0016	.0008	.0040	.0005	.39
%RSD	351.5	897.0	2.885	26.76	1.981
#1	-.0009	.0007	.1367	.0016	20.07
#2	.0025	.0000	.1390	.0026	20.05
#3	-.0004	-.0007	.1430	.0023	19.44
#4	.0018	.0011	.1321	.0018	19.73
#5	-.0008	-.0006	.1389	.0013	19.17

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.270
Stddev	.131
%RSD	1.2722

#1	10.122
#2	10.233
#3	10.208
#4	10.323
#5	10.466

2303

Analysis Report

02/26/99 14:28:29

page 1

Method: ISOU

Sample Name: s98t002604_s

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0001	.1049	.0057	51.37
Stddev	.0014	.0015	.0036	.0010	1.03
%RSD	156.8	1877.	3.443	17.00	1.999

#1	.0010	-.0022	.1059	.0063	51.29
#2	.0002	.0012	.1070	.0044	52.93
#3	-.0010	.0016	.1009	.0049	50.86
#4	.0014	-.0005	.1094	.0063	51.62
#5	.0027	.0004	.1016	.0066	50.16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.133
Stddev	.170
%RSD	1.6752

#1	10.198
#2	9.8898
#3	10.208
#4	10.040
#5	10.327

2304

Analysis Report

02/26/99 14:35:35

page 1

Method: ISOU

Sample Name: s98t002605

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:34 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	-.0018	.0002	-.0017	.1016
Stddev	.0012	.0013	.0011	.0008	.0069
%RSD	287.6	71.17	499.0	43.45	6.800

#1	-.0016	-.0014	-.0009	-.0012	.0948
#2	.0010	-.0006	.0000	-.0027	.1069
#3	.0014	-.0040	.0013	-.0008	.0950
#4	.0005	-.0012	-.0006	-.0017	.1013
#5	.0006	-.0017	.0014	-.0022	.1101

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.169
Stddev	.345
%RSD	3.3956

#1	10.247
#2	9.6647
#3	9.9970
#4	10.412
#5	10.527

2305

Analysis Report

02/26/99 14:38:44

page 1

Method: ISOU

Sample Name: s98t002605_d

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0003	-.0011	-.0021	.1055
Stddev	.0012	.0014	.0011	.0017	.0042
%RSD	142.3	468.2	105.7	81.27	3.984
#1	-.0017	-.0011	-.0014	-.0049	.1107
#2	-.0014	.0004	-.0008	-.0017	.1052
#3	-.0019	.0006	-.0029	-.0014	.0993
#4	-.0004	.0009	-.0002	-.0004	.1076
#5	.0011	-.0023	-.0001	-.0019	.1046

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.523
Stddev	.187
%RSD	1.7794

#1	10.483
#2	10.377
#3	10.772
#4	10.655
#5	10.328

2306

Analysis Report

02/26/99 14:40:47

page 1

Method: ISOU

Sample Name: s98t002606

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 14:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0035	-.0012	.0481	.0000	7.226
Stddev	.0002	.0016	.0027	.002	.103
%RSD	4.520	124.3	5.511	7420.	1.432
#1	.0037	-.0002	.0464	-.0019	7.219
#2	.0036	.0009	.0523	.0020	7.398
#3	.0035	-.0020	.0468	.0018	7.222
#4	.0035	-.0030	.0491	-.0011	7.153
#5	.0033	-.0018	.0458	-.0009	7.138

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.027
Stddev	.134
%RSD	1.2109

#1	10.964
#2	10.888
#3	11.106
#4	10.955
#5	11.219

2307

Analysis Report

02/26/99 14:42:42

page 1

Method: ISOU Sample Name: s98t002606_d Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 14:41 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0057	-.0011	.0558	.0004	8.327
Stddev	.0015	.0009	.0037	.0006	.212
%RSD	26.70	85.19	6.689	141.5	2.541

#1	.0051	-.0003	.0538	.0004	8.019
#2	.0061	-.0014	.0580	-.0004	8.586
#3	.0081	-.0003	.0574	.0012	8.318
#4	.0052	-.0008	.0596	.0002	8.444
#5	.0041	-.0025	.0504	.0006	8.268

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.228
Stddev	.248
%RSD	2.2109

#1	11.466
#2	11.061
#3	11.177
#4	10.937
#5	11.501

2308

Analysis Report

02/26/99 14:47:40

page 1

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso_u lcs 20 ppb

Run Time: 02/26/99 14:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0015	-.0007	.1475	.0012	21.17
Stddev	.0007	.0010	.0058	.0016	.73
%RSD	45.41	140.0	3.960	126.7	3.446
#1	-.0013	-.0019	.1405	-.0003	20.27
#2	-.0027	-.0014	.1513	.0034	21.84
#3	-.0012	.0000	.1553	-.0001	21.54
#4	-.0008	-.0010	.1448	.0021	21.70
#5	-.0017	.0006	.1456	.0010	20.49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.755
Stddev	.349
%RSD	2.9675

#1	12.239
#2	11.579
#3	11.407
#4	11.551
#5	11.998

2309

Analysis Report

02/26/99 15:05:28

page 1

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 15:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0008	-.0017	-.0015	-.0031	-.0024
Stddev	.0008	.0015	.0009	.0006	.0007
%RSD	99.66	90.88	57.23	19.92	30.35

#1	-.0016	.0004	-.0007	-.0036	-.0023
#2	-.0016	-.0018	-.0027	-.0021	-.0023
#3	.0001	-.0036	-.0014	-.0029	-.0032
#4	-.0007	-.0009	-.0020	-.0033	-.0028
#5	-.0001	-.0025	-.0007	-.0034	-.0013

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.791
Stddev	.012
%RSD	.09860

#1	11.772
#2	11.794
#3	11.798
#4	11.791
#5	11.801

2310

LABCORE Completed Worklist Report for Worklist# 27494

Analyst: bjg Instrument: ICPMS1

Book#: 42B57, 970119-3

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	MSPR-QC TC99	QC	0.0197	1.61e-02	81.726 %	Recovery
1	ICV	0	MSPR-QC SN117	QC	0.0077	6.81e-03	88.442 %	Recovery
1	ICV	0	MSPR-QC SN126	QC	n/a	<1.50e-4	%	Recovery
1	ICV	0	MSPR-QC CS133	QC	0.0200	2.15e-02	107.500 %	Recovery
1	ICV	0	MSPR-QC CS135	QC	n/a	3.25e-05	%	Recovery
1	ICV	0	MSPR-QC CS137	QC	n/a	5.69e-05	%	Recovery
2	ICB	0	MSPR-QC TC99	QC	1	<2.00e-6		ug/mL
2	ICB	0	MSPR-QC SN117	QC	1	<1.50e-4		ug/mL
2	ICB	0	MSPR-QC SN126	QC	1	<1.50e-4		ug/mL
2	ICB	0	MSPR-QC CS133	QC	1	<2.00e-6		ug/mL
2	ICB	0	MSPR-QC CS135	QC	1	<2.00e-6		ug/mL
2	ICB	0	MSPR-QC CS137	QC	1	<2.00e-6		ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 TC99-D1	LIQUID	N/A	4.001e+00	8.10e-002	ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 CS133-D1	LIQUID	N/A	5.164e+00	8.10e-002	ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 CS135-D1	LIQUID	N/A	1.442e+00	8.10e-002	ug/mL
3	SAMPLE	S98T002822	D D MSPR-D1 CS137-D1	LIQUID	N/A	1.150e+00	8.10e-002	ug/mL
4	DUP	S98T002822	D D MSPR-D1 TC99-D1	LIQUID	4.00e+00	4.12e+00	2.956	RPD
4	DUP	S98T002822	D D MSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
4	DUP	S98T002822	D D MSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
4	DUP	S98T002822	D D MSPR-D1 CS133-D1	LIQUID	5.16e+00	5.32e+00	3.053	RPD
4	DUP	S98T002822	D D MSPR-D1 CS135-D1	LIQUID	1.44e+00	1.49e+00	3.413	RPD
4	DUP	S98T002822	D D MSPR-D1 CS137-D1	LIQUID	1.15e+00	1.19e+00	3.419	RPD
5	SAMPLE	S98T002823	D D MSPR-D1 TC99-D1	LIQUID	N/A	4.386e+00	8.10e-002	ug/mL
5	SAMPLE	S98T002823	D D MSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
5	SAMPLE	S98T002823	D D MSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
5	SAMPLE	S98T002823	D D MSPR-D1 CS133-D1	LIQUID	N/A	5.771e+00	8.10e-002	ug/mL
5	SAMPLE	S98T002823	D D MSPR-D1 CS135-D1	LIQUID	N/A	1.624e+00	8.10e-002	ug/mL
5	SAMPLE	S98T002823	D D MSPR-D1 CS137-D1	LIQUID	N/A	1.397e+00	8.10e-002	ug/mL
6	DUP	S98T002823	D D MSPR-D1 TC99-D1	LIQUID	4.39e+00	4.43e+00	0.907	RPD
6	DUP	S98T002823	D D MSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
6	DUP	S98T002823	D D MSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
6	DUP	S98T002823	D D MSPR-D1 CS133-D1	LIQUID	5.77e+00	5.81e+00	0.691	RPD
6	DUP	S98T002823	D D MSPR-D1 CS135-D1	LIQUID	1.62e+00	1.63e+00	0.615	RPD
6	DUP	S98T002823	D D MSPR-D1 CS137-D1	LIQUID	1.40e+00	1.39e+00	0.717	RPD
7	SAMPLE	S98T002827	D D MSPR-D1 TC99-D1	LIQUID	N/A	4.192e+00	8.10e-002	ug/mL
7	SAMPLE	S98T002827	D D MSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
7	SAMPLE	S98T002827	D D MSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
7	SAMPLE	S98T002827	D D MSPR-D1 CS133-D1	LIQUID	N/A	1.023e+01	8.10e-002	ug/mL
7	SAMPLE	S98T002827	D D MSPR-D1 CS135-D1	LIQUID	N/A	3.730e+00	8.10e-002	ug/mL

12/09/98 15:25
A-0004-1

File # PR 0120 R. TAT

Page: 1

LABCORE Data Entry Template for Worklist# 27494

Analyst: B. G. G. G. Instrument: ICPMS1 _____ Book# 42657, 990119-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC	QC		
2 ICB			@MSPR-QC	QC		
3 SAMPLE	S98T002822	0 D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
4 DUP	S98T002822	0 D	@MSPR-D1	LIQUID		
5 SAMPLE	S98T002823	0 D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
6 DUP	S98T002823	0 D	@MSPR-D1	LIQUID		
7 SAMPLE	S98T002827	0 D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
8 DUP	S98T002827	0 D	@MSPR-D1	LIQUID		
9 SPK-POST	S98T002827	0 D	@MSPR-D1	LIQUID		
10 CCV			@MSPR-QC	QC		
11 CCB			@MSPR-QC	QC		

Final page for worklist # 27494

Analyst Signature B. G. G. G. Date 1-20-99

Analyst Signature _____ Date _____

Data Entry Comments:

SCANS = 982822, 982823, 982827

Validated + Uploaded by:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2313

1/21/99

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/20/99 13:46 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	117700.
Stddev	658.
%RSD	.5588

#1	117900.
#2	117200.
#3	118800.
#4	116900.
#5	117700.
#6	117600.
#7	117400.
#8	116700.
#9	118400.
#10	118100.

1-20-99

64-105

598708 2822, 23, 27

-1-40-1-025-10-1

FILE # PRO120A.TXT

W.L.H. 27494

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2314 TO 2328.

Analysis Report

Method: TC_SN_CS Sample Name: blank
Comment: background
Run Time: 01/20/99 13:48 Type: Unk

Operator: bjg

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0002	.0006	.0001	-.0005
Stddev	.0002	.0002	.0006	.0002	.0004
%RSD	84.26	99.05	113.9	159.9	70.13

#1	-.0006	.0000	.0000	.0000	-.0010
#2	-.0001	-.0001	.0011	.0001	.0000
#3	-.0001	-.0004	-.0003	.0000	-.0004
#4	-.0003	-.0001	.0010	.0000	-.0006
#5	-.0001	-.0005	.0008	.0003	-.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0052	-.0046	-.0074	-.0052
Stddev	.0005	.0007	.0010	.0002
%RSD	10.09	15.36	12.97	4.742

#1	-.0058	-.0054	-.0076	-.0056
#2	-.0048	-.0044	-.0076	-.0053
#3	-.0056	-.0041	-.0089	-.0050
#4	-.0049	-.0038	-.0063	-.0050
#5	-.0046	-.0051	-.0068	-.0053

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.101
Stddev	.067
%RSD	.54988

#1	12.120
#2	12.060
#3	12.081
#4	12.038
#5	12.208

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/20/99 13:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0057	-.0004	6.808	.0003	.0002
Stddev	.0010	.0002	.445	.0002	.0014
%RSD	18.14	59.08	6.529	77.72	836.1

#1	.0061	-.0007	6.610	.0003	-.0016
#2	.0066	-.0003	6.677	.0000	.0001
#3	.0043	-.0005	6.256	.0006	.0005
#4	.0048	-.0002	7.109	.0005	.0023
#5	.0065	-.0002	7.390	.0003	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0041	-.0085	-.0164	-.0120
Stddev	.0005	.0006	.0005	.0001
%RSD	11.48	7.220	3.138	.9519

#1	-.0048	-.0086	-.0167	-.0121
#2	-.0042	-.0091	-.0170	-.0121
#3	-.0041	-.0085	-.0163	-.0121
#4	-.0035	-.0075	-.0164	-.0119
#5	-.0040	-.0087	-.0157	-.0119

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.262
Stddev	.453
%RSD	4.0188

#1	11.450
#2	11.306
#3	11.884
#4	10.961
#5	10.710

Analysis Report

01/20/99 13:58:15

page 1

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/20/99 13:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.08	.0065	1.425	.0001	-.0012
Stddev	.27	.0004	.092	.0003	.0009
%RSD	1.654	6.637	6.444	323.2	77.82

#1	16.08	.0069	1.567	.0001	-.0010
#2	16.38	.0069	1.423	.0002	-.0024
#3	15.76	.0061	1.389	-.0002	.0002
#4	16.30	.0060	1.431	-.0002	-.0014
#5	15.87	.0067	1.315	.0005	-.0014

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.53	.0325	.0569	.0388
Stddev	.26	.0009	.0020	.0011
%RSD	1.198	2.643	3.429	2.821

#1	21.81	.0334	.0589	.0398
#2	21.25	.0314	.0590	.0378
#3	21.27	.0333	.0548	.0376
#4	21.73	.0321	.0555	.0400
#5	21.57	.0320	.0562	.0387

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.545
Stddev	.080
%RSD	.69709

#1	11.478
#2	11.568
#3	11.623
#4	11.444
#5	11.612

2317

Analysis Report

01/20/99 14:04:51

page 1

Method: TC_SN_CS Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/20/99 14:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	-.0004	.0004	.0000	-.0008
Stddev	.000	.0001	.0008	.0004	.0008
%RSD	3911.	25.84	228.4	1398.	100.0

#1	-.0001	-.0003	.0003	.0007	.0002
#2	.0003	-.0004	-.0004	-.0002	-.0002
#3	.0002	-.0005	-.0005	-.0001	-.0011
#4	-.0001	-.0004	.0011	.0000	-.0011
#5	-.0002	-.0005	.0013	-.0003	-.0016

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0027	-.0035	-.0075	-.0053
Stddev	.0005	.0004	.0009	.0002
%RSD	16.76	11.92	12.04	3.387

#1	-.0032	-.0041	-.0076	-.0053
#2	-.0027	-.0035	-.0060	-.0055
#3	-.0025	-.0030	-.0082	-.0056
#4	-.0021	-.0033	-.0082	-.0051
#5	-.0031	-.0033	-.0073	-.0053

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.977
Stddev	.095
%RSD	.79051

#1	11.987
#2	12.007
#3	12.034
#4	12.044
#5	11.812

2318

Analysis Report

01/20/99 14:07:55

page 1

Method: TC_SN_CS Sample Name: s98t002822

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0988	.1258	.0333	.0008	-.0002
Stddev	.0100	.0089	.0046	.0007	.0004
%RSD	10.16	7.080	13.72	86.79	162.4

#1	.0832	.1110	.0255	.0004	-.0007
#2	.1110	.1340	.0373	.0017	.0001
#3	.0992	.1297	.0337	.0012	-.0005
#4	.0984	.1297	.0340	.0004	-.0001
#5	.1022	.1246	.0358	.0001	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1275	.0356	.0284	-.0125
Stddev	.0061	.0028	.0031	.0001
%RSD	4.763	7.782	11.04	1.104

#1	.1171	.0312	.0234	-.0126
#2	.1327	.0377	.0312	-.0122
#3	.1275	.0381	.0309	-.0125
#4	.1302	.0350	.0284	-.0125
#5	.1298	.0361	.0279	-.0125

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.086
Stddev	.892
%RSD	7.3778

#1	13.604
#2	11.249
#3	11.848
#4	11.761
#5	11.968

2319

Analysis Report

01/20/99 14:12:40

page 1

Method: TC_SN_CS Sample Name: s98t002822_d

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:10 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1018	.1329	.0337	.0014	-.0008
Stddev	.0037	.0035	.0021	.0004	.0011
%RSD	3.644	2.634	6.236	30.41	140.2

#1	.1024	.1360	.0348	.0011	-.0005
#2	.1009	.1336	.0367	.0017	-.0008
#3	.1061	.1363	.0331	.0017	.0006
#4	.0961	.1291	.0311	.0007	-.0008
#5	.1035	.1293	.0330	.0016	-.0024

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1313	.0367	.0294	-.0126
Stddev	.0013	.0019	.0011	.0001
%RSD	.9545	5.058	3.850	.8581

#1	.1304	.0395	.0301	-.0124
#2	.1295	.0354	.0308	-.0125
#3	.1322	.0376	.0297	-.0126
#4	.1319	.0350	.0287	-.0126
#5	.1323	.0358	.0279	-.0126

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.805
Stddev	.212
%RSD	1.7935

#1	11.860
#2	11.657
#3	11.524
#4	12.047
#5	11.936

2320

Analysis Report

01/20/99 14:15:28

page 1

Method: TC_SN_CS Sample Name: s98t002823

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1083	.1329	.0363	.0011	.0002
Stddev	.0025	.0032	.0010	.0004	.0006
%RSD	2.333	2.379	2.664	31.69	302.1

#1	.1091	.1356	.0370	.0010	.0008
#2	.1049	.1308	.0353	.0015	.0004
#3	.1118	.1293	.0370	.0006	-.0006
#4	.1072	.1320	.0372	.0012	.0006
#5	.1083	.1367	.0353	.0012	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1425	.0401	.0345	-.0125
Stddev	.0043	.0028	.0016	.0001
%RSD	2.998	7.080	4.704	1.006

#1	.1465	.0391	.0356	-.0126
#2	.1477	.0386	.0348	-.0123
#3	.1383	.0404	.0319	-.0125
#4	.1396	.0448	.0341	-.0125
#5	.1405	.0375	.0360	-.0126

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.580
Stddev	.065
%RSD	.55963

#1	11.639
#2	11.520
#3	11.661
#4	11.534
#5	11.547

2321

Analysis Report

01/20/99 14:18:51

page 1

Method: TC_SN_CS Sample Name: s98t002823_d Operator: bjg
 Comment: 40501 df
 Run Time: 01/20/99 14:16 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1095	.1314	.0387	.0012	-.0009
Stddev	.0037	.0053	.0015	.0006	.0007
%RSD	3.374	4.024	3.808	49.66	77.35
#1	.1052	.1246	.0371	.0013	-.0008
#2	.1062	.1286	.0375	.0016	-.0006
#3	.1138	.1362	.0407	.0004	-.0013
#4	.1119	.1372	.0396	.0008	-.0017
#5	.1102	.1306	.0387	.0019	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1434	.0403	.0343	-.0126
Stddev	.0021	.0016	.0020	.0001
%RSD	1.464	3.879	5.929	.6212
#1	.1446	.0412	.0370	-.0126
#2	.1421	.0396	.0334	-.0125
#3	.1417	.0403	.0353	-.0126
#4	.1422	.0382	.0340	-.0126
#5	.1466	.0423	.0316	-.0125

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.462
Stddev	.152
%RSD	1.3265

#1	11.536
#2	11.687
#3	11.367
#4	11.303
#5	11.415

2322

Analysis Report

01/20/99 14:21:40

page 1

Method: TC_SN_CS Sample Name: s98t002827

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1035	.1311	.0372	.0011	-.0004
Stddev	.0027	.0037	.0012	.0004	.0005
%RSD	2.628	2.841	3.222	36.57	129.7

#1	.1057	.1339	.0377	.0011	.0002
#2	.1029	.1352	.0381	.0018	-.0006
#3	.1020	.1263	.0380	.0010	-.0010
#4	.1001	.1283	.0371	.0007	-.0004
#5	.1068	.1318	.0352	.0011	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2525	.0921	.0674	-.0125
Stddev	.0035	.0028	.0029	.0000
%RSD	1.402	3.055	4.282	.2885

#1	.2523	.0932	.0664	-.0124
#2	.2549	.0912	.0633	-.0124
#3	.2474	.0889	.0685	-.0125
#4	.2566	.0964	.0678	-.0125
#5	.2513	.0912	.0711	-.0125

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.575
Stddev	.068
%RSD	.58538

#1	11.520
#2	11.526
#3	11.533
#4	11.630
#5	11.665

2323

Analysis Report

01/20/99 14:25:53

page 1

Method: TC_SN_CS Sample Name: s98t002827_d

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1060	.1314	.0389	.0014	.0001
Stddev	.0030	.0043	.0014	.0004	.0007
%RSD	2.798	3.275	3.569	26.56	606.2

#1	.1059	.1284	.0397	.0009	.0005
#2	.1092	.1309	.0407	.0017	-.0002
#3	.1058	.1371	.0389	.0012	.0010
#4	.1079	.1341	.0384	.0018	-.0003
#5	.1013	.1264	.0370	.0013	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2568	.0904	.0699	-.0125
Stddev	.0055	.0031	.0012	.0001
%RSD	2.139	3.430	1.690	.5396

#1	.2511	.0857	.0704	-.0125
#2	.2652	.0913	.0712	-.0124
#3	.2535	.0929	.0706	-.0125
#4	.2552	.0932	.0686	-.0125
#5	.2590	.0891	.0687	-.0126

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.879
Stddev	.240
%RSD	2.0214

#1	12.158
#2	11.504
#3	11.851
#4	11.900
#5	11.985

2324

Method: TC_SN_CS Sample Name: s98t002827_a

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 14:26 Type: Unk

Mode: CONC

Corr-Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.47	.1317	4.448	.0013	-.0003
Stddev	.77	.0070	.238	.0003	.0007
%RSD	4.400	5.286	5.358	22.73	223.8

#1	16.77	.1310	4.354	.0012	.0002
#2	17.69	.1303	4.567	.0010	.0002
#3	17.96	.1379	4.598	.0010	-.0007
#4	18.35	.1383	4.648	.0016	-.0014
#5	16.57	.1212	4.072	.0016	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.51	.0909	.0710	-.0113
Stddev	1.05	.0043	.0060	.0001
%RSD	4.888	4.720	8.476	.7732

#1	21.00	.0878	.0696	-.0112
#2	21.42	.0925	.0736	-.0113
#3	22.27	.0931	.0721	-.0114
#4	22.76	.0958	.0779	-.0114
#5	20.10	.0852	.0617	-.0113

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.531
Stddev	.491
%RSD	4.2612

#1	11.641
#2	11.507
#3	11.195
#4	11.020
#5	12.292

Analysis Report

01/20/99 14:35:13

page 1

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/20/99 14:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0085	-.0005	6.748	.0001	.0002
Stddev	.0003	.0003	.171	.0005	.0013
%RSD	3.664	54.69	2.528	604.0	838.9

#1	.0090	-.0008	6.740	-.0006	-.0010
#2	.0086	-.0005	6.932	.0006	.0016
#3	.0083	-.0005	6.482	.0005	-.0007
#4	.0083	-.0001	6.855	-.0003	-.0007
#5	.0082	-.0007	6.728	.0003	.0017

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0044	-.0095	-.0172	-.0124
Stddev	.0015	.0014	.0016	.0007
%RSD	33.65	14.34	9.480	5.777

#1	-.0044	-.0101	-.0175	-.0120
#2	-.0056	-.0085	-.0157	-.0119
#3	-.0057	-.0089	-.0168	-.0122
#4	-.0046	-.0084	-.0162	-.0121
#5	-.0020	-.0116	-.0199	-.0137

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.654
Stddev	.388
%RSD	3.3260

#1	11.750
#2	11.231
#3	11.900
#4	11.278
#5	12.113

2326

Analysis Report

01/20/99 14:39:27

page 1

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/20/99 14:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.03	.0063	1.410	.0001	.0002
Stddev	.20	.0007	.074	.0003	.0006
%RSD	1.255	10.95	5.257	465.6	304.8

#1	16.20	.0066	1.283	.0003	.0011
#2	16.04	.0060	1.440	.0001	.0007
#3	16.26	.0053	1.410	.0000	.0001
#4	15.81	.0070	1.468	.0000	.0003
#5	15.85	.0068	1.451	.0006	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.36	.0329	.0577	.0391
Stddev	.32	.0020	.0014	.0006
%RSD	1.520	5.996	2.474	1.458

#1	20.87	.0320	.0581	.0388
#2	21.74	.0341	.0567	.0393
#3	21.32	.0333	.0568	.0400
#4	21.54	.0352	.0569	.0390
#5	21.33	.0301	.0600	.0385

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.250
Stddev	.066
%RSD	.58658

#1	11.335
#2	11.176
#3	11.226
#4	11.214
#5	11.302

2327

Analysis Report

01/20/99 14:46:46

page 1

Method: TC_SN_CS Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/20/99 14:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	-.0005	-.0001	.0001	-.0009
Stddev	.0004	.0001	.0009	.0003	.0008
%RSD	744.3	26.56	1044.	626.4	87.24

#1	.0000	-.0005	.0007	.0002	-.0006
#2	.0005	-.0003	-.0003	-.0002	-.0004
#3	-.0006	-.0006	-.0012	-.0004	-.0001
#4	.0000	-.0005	-.0005	.0004	-.0016
#5	.0003	-.0005	.0009	.0003	-.0020

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0050	-.0040	-.0065	-.0052
Stddev	.0007	.0009	.0005	.0004
%RSD	13.94	21.67	8.408	7.554

#1	-.0056	-.0049	-.0072	-.0055
#2	-.0044	-.0029	-.0070	-.0046
#3	-.0055	-.0049	-.0058	-.0056
#4	-.0055	-.0036	-.0062	-.0053
#5	-.0041	-.0039	-.0065	-.0050

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.584
Stddev	.451
%RSD	3.8932

#1	12.246
#2	11.013
#3	11.727
#4	11.489
#5	11.445

2328

LABCORE Completed Worklist Report for Worklist# 27495

Analyst: **kjt** *KJT*

Instrument: ICPMS1

Book#: 990104-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	QMSU-QC U235	QC	0.000143	1.46e-04	102.098 % Recovery	
1 ICV	0	QMSU-QC U238	QC	0.020	2.10e-02	105.000 % Recovery	
2 ICB	0	QMSU-QC U233	QC	1	<5.07e-5		ug/mL
2 ICB	0	QMSU-QC U234	QC	1	<4.10e-5		ug/mL
2 ICB	0	QMSU-QC U235	QC	1	<4.55e-5		ug/mL
2 ICB	0	QMSU-QC U236	QC	1	<3.57e-5		ug/mL
2 ICB	0	QMSU-QC U238	QC	1	<5.86e-5		ug/mL
3 SAMPLE	S98T002822 0 D	QMSU-D1 U233-D1	LIQUID	<u>N/A</u> <	2.053e+00	2.053	ug/mL
3 SAMPLE	S98T002822 0 D	QMSU-D1 U234-D1	LIQUID	<u>N/A</u> <	1.659e+00	1.659	ug/mL
3 SAMPLE	S98T002822 0 D	QMSU-D1 U235-D1	LIQUID	<u>N/A</u> <	1.845e+00	1.845	ug/mL
3 SAMPLE	S98T002822 0 D	QMSU-D1 U236-D1	LIQUID	<u>N/A</u> <	1.448e+00	1.448	ug/mL
3 SAMPLE	S98T002822 0 D	QMSU-D1 U238-D1	LIQUID	<u>N/A</u> <	2.373e+00	2.373	ug/mL
4 DUP	S98T002822 0 D	QMSU-D1 U233-D1	LIQUID	<2.05e0	<2.05e0		RPD
4 DUP	S98T002822 0 D	QMSU-D1 U234-D1	LIQUID	<1.66e0	<1.66e0		RPD
4 DUP	S98T002822 0 D	QMSU-D1 U235-D1	LIQUID	<1.84e0	<1.84e0		RPD
4 DUP	S98T002822 0 D	QMSU-D1 U236-D1	LIQUID	<1.45e0	<1.45e0		RPD
4 DUP	S98T002822 0 D	QMSU-D1 U238-D1	LIQUID	<2.37e0	<2.37e0		RPD
5 SAMPLE	S98T002823 0 D	QMSU-D1 U233-D1	LIQUID	<u>N/A</u> <	2.053e+00	2.053	ug/mL
5 SAMPLE	S98T002823 0 D	QMSU-D1 U234-D1	LIQUID	<u>N/A</u> <	1.659e+00	1.659	ug/mL
5 SAMPLE	S98T002823 0 D	QMSU-D1 U235-D1	LIQUID	<u>N/A</u> <	1.845e+00	1.845	ug/mL
5 SAMPLE	S98T002823 0 D	QMSU-D1 U236-D1	LIQUID	<u>N/A</u> <	1.448e+00	1.448	ug/mL
5 SAMPLE	S98T002823 0 D	QMSU-D1 U238-D1	LIQUID	<u>N/A</u> <	2.373e+00	2.373	ug/mL
6 DUP	S98T002823 0 D	QMSU-D1 U233-D1	LIQUID	<2.05e0	<2.05e0		RPD
6 DUP	S98T002823 0 D	QMSU-D1 U234-D1	LIQUID	<1.66e0	<1.66e0		RPD
6 DUP	S98T002823 0 D	QMSU-D1 U235-D1	LIQUID	<1.84e0	<1.84e0		RPD
6 DUP	S98T002823 0 D	QMSU-D1 U236-D1	LIQUID	<1.45e0	<1.45e0		RPD
6 DUP	S98T002823 0 D	QMSU-D1 U238-D1	LIQUID	<2.37e0	<2.37e0		RPD
7 SAMPLE	S98T002827 0 D	QMSU-D1 U233-D1	LIQUID	<u>N/A</u> <	2.053e+00	2.053	ug/mL
7 SAMPLE	S98T002827 0 D	QMSU-D1 U234-D1	LIQUID	<u>N/A</u> <	1.659e+00	1.659	ug/mL
7 SAMPLE	S98T002827 0 D	QMSU-D1 U235-D1	LIQUID	<u>N/A</u> <	1.845e+00	1.845	ug/mL
7 SAMPLE	S98T002827 0 D	QMSU-D1 U236-D1	LIQUID	<u>N/A</u> <	1.448e+00	1.448	ug/mL
7 SAMPLE	S98T002827 0 D	QMSU-D1 U238-D1	LIQUID	<u>N/A</u> <	2.373e+00	2.373	ug/mL
8 DUP	S98T002827 0 D	QMSU-D1 U233-D1	LIQUID	<2.05e0	<2.05e0		RPD
8 DUP	S98T002827 0 D	QMSU-D1 U234-D1	LIQUID	<1.66e0	<1.66e0		RPD
8 DUP	S98T002827 0 D	QMSU-D1 U235-D1	LIQUID	<1.84e0	<1.84e0		RPD
8 DUP	S98T002827 0 D	QMSU-D1 U236-D1	LIQUID	<1.45e0	<1.45e0		RPD
8 DUP	S98T002827 0 D	QMSU-D1 U238-D1	LIQUID	<2.37e0	<2.37e0		RPD
9 SPK-POST	S98T002827 0 D	QMSU-D1 U233-D1	LIQUID	n/a	-1.76e-6		% Recovery
9 SPK-POST	S98T002827 0 D	QMSU-D1 U234-D1	LIQUID	n/a	3.78e-07		% Recovery
9 SPK-POST	S98T002827 0 D	QMSU-D1 U235-D1	LIQUID	n/a	1.34e-04		% Recovery
9 SPK-POST	S98T002827 0 D	QMSU-D1 U236-D1	LIQUID	n/a	3.58e-06		% Recovery

LABCORE Completed Worklist Report for Worklist# 27495

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 SPK-POST	S98T002827	0 D	QMSU-D1 U238-D1	LIQUID	0.020	2.00e-02	100.000	% Recovery
10 CCV	0		QMSU-QC U235	QC	0.000143	1.34e-04	93.706	% Recovery
10 CCV	0		QMSU-QC U238	QC	0.020	2.03e-02	101.500	% Recovery
11 CCB	0		QMSU-QC U233	QC	1	<5.07e-5		ug/mL
11 CCB	0		QMSU-QC U234	QC	1	<4.10e-5		ug/mL
11 CCB	0		QMSU-QC U235	QC	1	<4.55e-5		ug/mL
11 CCB	0		QMSU-QC U236	QC	1	<3.57e-5		ug/mL
11 CCB	0		QMSU-QC U238	QC	1	<5.86e-5		ug/mL

Final page for worklist# 27495

_____
Analyst Signature   Date

_____
Analyst Signature   Date

Saul H. Pung   1-12-99
_____
Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 27495

Analyst: KST Instrument: ICPMS1 _____ Book# 990104-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A Test	Matrix	Group#	Project
1 ICV		@MSU-QC	QC		
2 ICB		@MSU-QC	QC		
3 SAMPLE	S98T002822 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1					
4 DUP	S98T002822 0 D	@MSU-D1	LIQUID		
5 SAMPLE	S98T002823 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1					
6 DUP	S98T002823 0 D	@MSU-D1	LIQUID		
7 SAMPLE	S98T002827 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1					
8 DUP	S98T002827 0 D	@MSU-D1	LIQUID		
9 SPK-POST	S98T002827 0 D	@MSU-D1	LIQUID		
10 CCV		@MSU-QC	QC		
11 CCB		@MSU-QC	QC		

Final page for worklist # 27495

K. S. Thomas
Analyst Signature Date

_____
Analyst Signature Date

S98T002822, 23, 27 .100-10 - .025-10
file 01116.txt

Data Entry Comments:

Uploaded + Validated by:
Sam M. Pang
1-12-99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2331

Method: MS_TUNE   Sample Name: ms-tune   Operator: kjt
Comment:
Run Time: 01/11/99 13:37 Type: Unk   Mode: INT   Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	111000.
Stddev	1408.
%RSD	1.268

#1	111000.
#2	111300.
#3	108300.
#4	111700.
#5	112500.
#6	110300.
#7	110800.
#8	113400.
#9	110000.
#10	111000.

K5 thermidike 1-11-98
WL 27495 file 001116.txt
3981002822 01-10-.025-10
BY 105 2823 01-10-.025-10
2827 01-10-.025-10
2827-a +.020 10ppm U

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2332 TO 2352

Method: ISO_U Sample Name: iso-u

Operator: kjt

Comment: iso-u lcs

Run Time: 01/11/99 13:52 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.334	7.634	185.4	9.147	22640.
Stddev	.585	.876	6.5	.893	1087.
%RSD	10.97	11.48	3.515	9.762	4.802

#1	4.592	6.853	174.4	9.538	20880.
#2	5.927	8.130	184.4	10.26	22940.
#3	5.624	8.936	188.7	8.551	23450.
#4	5.696	7.120	190.3	9.414	23550.
#5	4.830	7.131	189.0	7.974	22390.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5979
Stddev	.2849
%RSD	4.3179

#1	7.0773
#2	6.5803
#3	6.4904
#4	6.3203
#5	6.5212

Method: ISO_U Sample Name: iso-u

Operator: kjt

Comment: iso-u lcs

Run Time: 01/11/99 13:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.334	7.634	.7490	9.147	104.1
Stddev	.585	.876	.0272	.893	5.0
%RSD	10.97	11.48	3.636	9.762	4.804
#1	4.592	6.853	.7032	9.538	96.05
#2	5.927	8.130	.7450	10.26	105.5
#3	5.624	8.936	.7631	8.551	107.9
#4	5.696	7.120	.7698	9.414	108.3
#5	4.830	7.131	.7641	7.974	103.0

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5979
Stddev	.2849
%RSD	4.3179

#1	7.0773
#2	6.5803
#3	6.4904
#4	6.3203
#5	6.5212

Method: ISO_U Sample Name: blank

Operator: kjt

Comment: background

Run Time: 01/11/99 13:59 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.777	5.166	5.277	5.178	5.635
Stddev	.830	.671	.746	.585	.959
%RSD	17.37	12.98	14.13	11.30	17.02
#1	3.363	4.580	5.725	4.437	5.009
#2	5.360	5.948	6.242	5.287	5.581
#3	4.817	5.839	4.671	6.058	5.715
#4	4.947	4.817	4.426	5.012	4.426
#5	5.397	4.647	5.322	5.097	5.446

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9998
Stddev	.3975
%RSD	5.6786

#1	6.9870
#2	6.8091
#3	6.8503
#4	7.6816
#5	6.6707

Method: ISO_U   Sample Name: blank
Comment: background
Run Time: 01/11/99 13:59 Type: Unk

Operator: kjt
Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.777	5.166	-.0039	5.178	-.0029
Stddev	.830	.671	.0031	.585	.0044
%RSD	17.37	12.98	79.38	11.30	151.3

#1	3.363	4.580	-.0021	4.437	-.0058
#2	5.360	5.948	.0001	5.287	-.0032
#3	4.817	5.839	-.0065	6.058	.0021
#4	4.947	4.817	-.0075	5.012	-.0085
#5	5.397	4.647	-.0037	5.097	.0008

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9998
Stddev	.3975
%RSD	5.6786

#1	6.9870
#2	6.8091
#3	6.8503
#4	7.6816
#5	6.6707

Method: ISO_U Sample Name: icv

Operator: kjt

Comment: 20 ppb u

Run Time: 01/11/99 14:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.571	5.703	41.20	5.893	4580.
Stddev	.936	1.655	1.48	.600	91.
%RSD	16.81	29.02	3.586	10.17	1.986
#1	4.428	7.675	38.74	6.273	4474.
#2	7.036	6.424	41.53	5.353	4537.
#3	5.346	3.437	41.09	6.416	4662.
#4	5.525	6.303	42.17	5.136	4687.
#5	5.521	4.677	42.48	6.287	4538.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5612
Stddev	.1289
%RSD	1.9649

#1	6.7751
#2	6.5378
#3	6.5464
#4	6.4258
#5	6.5209

Method: ISO_U Sample Name: icv

Operator: kjt

Comment: 20 ppb u

Run Time: 01/11/99 14:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.571	5.703	.1463	5.893	21.04
Stddev	.936	1.655	.0062	.600	.42
%RSD	16.81	29.02	4.223	10.17	1.989
#1	4.428	7.675	.1360	6.273	20.56
#2	7.036	6.424	.1476	5.353	20.85
#3	5.346	3.437	.1458	6.416	21.42
#4	5.525	6.303	.1503	5.136	21.54
#5	5.521	4.677	.1516	6.287	20.85

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5612
Stddev	.1289
%RSD	1.9649

#1	6.7751
#2	6.5378
#3	6.5464
#4	6.4258
#5	6.5209

Analysis Report

Method: ISO_U Sample Name: icb

Operator: kjt

Comment: 2% hno3

Run Time: 01/11/99 14:11 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.402	5.325	4.900	5.017	5.442
Stddev	.689	.523	.441	.853	.737
%RSD	12.76	9.832	8.999	16.96	13.53

#1	4.240	4.657	5.144	5.283	4.866
#2	5.663	6.021	5.448	5.520	6.093
#3	5.410	5.044	4.533	3.875	5.191
#4	5.641	5.284	4.998	4.427	4.712
#5	6.055	5.617	4.377	5.982	6.347

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9728
Stddev	.1425
%RSD	2.0437

#1	7.1927
#2	6.9751
#3	6.8393
#4	7.0029
#5	6.8538

Method: ISO_U Sample Name: icb

Operator: kjt

Comment: 2% hno3

Run Time: 01/11/99 14:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.402	5.325	-.0055	5.017	-.0038
Stddev	.689	.523	.0018	.853	.0034
%RSD	12.76	9.832	33.49	16.99	88.99
#1	4.240	4.657	-.0045	5.283	-.0065
#2	5.663	6.021	-.0032	5.520	-.0008
#3	5.410	5.044	-.0070	3.875	-.0050
#4	5.641	5.284	-.0051	4.427	-.0072
#5	6.055	5.617	-.0077	5.982	.0004

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9728
Stddev	.1425
%RSD	2.0437

#1	7.1927
#2	6.9751
#3	6.8393
#4	7.0029
#5	6.8538

Method: ISO_U Sample Name: s98t002822

Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.400	4.813	5.100	5.337	9.295
Stddev	.513	.448	.578	.545	.758
%RSD	9.502	9.300	11.34	10.21	8.158

#1	5.371	5.224	4.268	5.077	8.903
#2	5.472	4.437	5.694	5.546	9.244
#3	4.822	4.294	5.425	5.048	10.10
#4	5.133	5.284	4.756	6.190	9.964
#5	6.201	4.823	5.359	4.823	8.268

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6696
Stddev	.1078
%RSD	1.6161

#1	6.7953
#2	6.7611
#3	6.6365
#4	6.6238
#5	6.5313

Method: ISO_U   Sample Name: s98t002822   Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:13 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.400	4.813	-.0047	5.337	.0139
Stddev	.513	.448	.0024	.545	.0035
%RSD	9.502	9.300	51.79	10.21	25.06

#1	5.371	5.224	-.0081	5.077	.0121
#2	5.472	4.437	-.0022	5.546	.0137
#3	4.822	4.294	-.0033	5.048	.0176
#4	5.133	5.284	-.0061	6.190	.0170
#5	6.201	4.823	-.0036	4.823	.0092

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6696
Stddev	.1078
%RSD	1.6161

#1	6.7953
#2	6.7611
#3	6.6365
#4	6.6238
#5	6.5313

HNF-1663 REV. 0

Method: ISO_U Sample Name: s98t002822_d Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:15 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.107	5.141	5.435	5.377	9.491
Stddev	.543	.793	.527	.722	.954
%RSD	10.64	15.43	9.694	13.43	10.05
#1	4.413	4.118	5.148	4.854	10.81
#2	5.651	4.658	5.804	5.956	10.00
#3	5.239	5.087	6.150	6.150	9.415
#4	4.676	5.826	5.212	4.446	8.815
#5	5.554	6.017	4.860	5.477	8.408

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5872
Stddev	.1241
%RSD	1.8843

#1	6.7988
#2	6.5476
#3	6.5850
#4	6.5230
#5	6.4815

HNF-1663 REV.0

Method: ISO_U

Sample Name: s98t002822_d

Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.107	5.141	-.0033	5.377	.0148
Stddev	.543	.793	.0022	.722	.0044
%RSD	10.64	15.43	67.40	13.43	29.61
#1	4.413	4.118	-.0045	4.854	.0209
#2	5.651	4.658	-.0017	5.956	.0172
#3	5.239	5.087	-.0003	6.150	.0145
#4	4.676	5.826	-.0042	4.446	.0117
#5	5.554	6.017	-.0057	5.477	.0098

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5872
Stddev	.1241
%RSD	1.8843

#1	6.7988
#2	6.5476
#3	6.5850
#4	6.5230
#5	6.4815

Method: ISO_U Sample Name: s98t002823

Operator: xjt

Comment: 40501 df

Run Time: 01/11/99 14:17 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.038	5.342	4.789	5.297	9.443
Stddev	.530	.637	.507	1.230	.763
%RSD	10.52	11.92	10.59	23.22	8.077
#1	5.556	5.861	5.252	6.013	9.134
#2	5.081	4.399	4.854	5.537	9.253
#3	4.166	5.454	4.772	3.409	10.30
#4	5.342	5.049	3.951	4.903	10.10
#5	5.042	5.946	5.118	6.623	8.429

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6477
Stddev	.1070
%RSD	1.6101

#1	6.5689
#2	6.5926
#3	6.6005
#4	6.8330
#5	6.6437

Method: ISO_U Sample Name: s98t002823

Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:17 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.038	5.342	-.0060	5.297	.0146
Stddev	.530	.637	.0021	1.230	.0035
%RSD	10.52	11.92	35.53	23.22	24.03
#1	5.556	5.861	-.0040	6.013	.0132
#2	5.081	4.399	-.0057	5.537	.0137
#3	4.166	5.454	-.0060	3.409	.0186
#4	5.342	5.049	-.0095	4.903	.0176
#5	5.042	5.946	-.0046	6.623	.0099

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6477
Stddev	.1070
%RSD	1.6101

#1	6.5689
#2	6.5926
#3	6.6005
#4	6.8330
#5	6.6437

Method: ISO_U

Sample Name: s98t002823_d

Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:19 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.930	5.461	5.123	5.588	6.639
Stddev	.495	1.044	.652	.637	.391
%RSD	10.04	19.12	12.73	11.40	4.528
#1	4.516	7.150	4.967	5.268	6.429
#2	4.585	4.974	6.217	4.818	5.170
#3	4.787	4.937	4.488	5.461	6.901
#4	5.743	5.743	4.848	6.489	6.503
#5	5.019	4.503	5.093	5.905	8.194

Int. Std.

Ir[193]

Line	193/pulse
Units	Cts/S
Avg	6.6478
Stddev	.1285
%RSD	1.9331

#1	6.6435
#2	6.4339
#3	6.6843
#4	6.7036
#5	6.7736

Method: ISO_U

Sample Name: s98t002823_d

Operator: xjt

Comment: 40501 df

Run Time: 01/11/99 14:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.930	5.461	- .0046	5.588	.0109
Stddev	.495	1.044	.0027	.637	.0018
%RSD	10.04	19.12	59.62	11.40	16.50
#1	4.516	7.150	- .0052	5.268	.0099
#2	4.585	4.974	.0000	4.818	.0133
#3	4.787	4.937	- .0072	5.461	.0121
#4	5.743	5.743	- .0057	6.489	.0103
#5	5.019	4.503	- .0047	5.905	.0089

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6478
Stddev	.1285
%RSD	1.9331

#1	6.6435
#2	6.4339
#3	6.6843
#4	6.7036
#5	6.7736

HNF-1663 REV.0

Method: ISO_U

Sample Name: s98t002827

Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:21 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.649	4.640	4.989	4.677	7.514
Stddev	.655	.495	.551	.216	.307
%RSD	14.09	10.66	11.05	4.630	4.091
#1	4.163	3.797	4.309	4.601	7.303
#2	4.900	4.900	5.589	4.517	7.885
#3	4.582	4.657	4.507	4.807	7.812
#4	5.627	4.779	5.242	4.471	7.246
#5	3.975	5.066	5.300	4.988	7.326

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5873
Stddev	.1696
%RSD	2.5743

#1	6.8467
#2	6.5312
#3	6.6567
#4	6.4862
#5	6.4154

Method: ISO_U   Sample Name: s98t002827   Operator: kjt

Comment: 40501 df

Run Time: 01/11/99 14:21 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.649	4.640	-.0051	4.677	.0057
Stddev	.655	.495	.0023	.216	.0014
%RSD	14.09	10.66	44.91	4.630	24.70
#1	4.163	3.797	-.0080	4.601	.0048
#2	4.900	4.900	-.0026	4.517	.0074
#3	4.582	4.657	-.0071	4.807	.0071
#4	5.627	4.779	-.0041	4.471	.0045
#5	3.975	5.066	-.0038	4.988	.0049

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5873
Stddev	.1696
%RSD	2.5743

#1	6.8467
#2	6.5312
#3	6.6567
#4	6.4862
#5	6.4154

Method: ISO_U   Sample Name: s98t002827_d   Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:23 Type: Unk   Mode: IR   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.524	5.505	4.945	5.163	7.661
Stddev	.470	.434	.305	.635	.896
%RSD	10.38	7.892	6.171	12.30	11.69
#1	5.010	6.167	5.010	6.013	8.865
#2	4.443	5.628	5.036	4.443	8.220
#3	4.783	5.007	4.409	4.633	7.174
#4	4.612	5.444	5.141	5.444	6.578
#5	3.772	5.281	5.130	5.281	7.469

Int. Std.   Ir[193]

Line	193/pulse
Units	Cts/S
Avg	6.6339
Stddev	.0992
%RSD	1.4948
#1	6.4864
#2	6.7517
#3	6.6907
#4	6.6133
#5	6.6272

Method: ISO_U   Sample Name: s98t002827_d   Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:23 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.524	5.505	-.0053	5.163	.0064
Stddev	.470	.434	.0013	.635	.0041
%RSD	10.38	7.892	24.01	12.30	64.36
#1	5.010	6.167	-.0050	6.013	.0119
#2	4.443	5.628	-.0049	4.443	.0090
#3	4.783	5.007	-.0076	4.633	.0042
#4	4.612	5.444	-.0045	5.444	.0014
#5	3.772	5.281	-.0045	5.281	.0055

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.6339
Stddev	.0992
%RSD	1.4948

#1	6.4864
#2	6.7517
#3	6.6907
#4	6.6133
#5	6.6272

Method: ISO_U   Sample Name: s98t002827_a   Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:25 Type: Unk   Mode: IR   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.343	5.259	38.33	6.059	4362.
Stddev	.551	.564	1.79	.464	83.
%RSD	12.68	10.73	4.664	7.653	1.901
#1	3.936	5.612	36.44	5.612	4354.
#2	4.386	5.824	36.31	6.471	4327.
#3	4.574	4.646	39.49	5.517	4304.
#4	3.709	4.654	39.78	6.472	4319.
#5	5.112	5.556	39.63	6.223	4507.

Int. Std.,   Ir[193]
Line   193/pulse
Units   Cts/S
Avg   6.8652
Stddev   .0740
%RSD   1.0781

#1	6.8599
#2	6.9539
#3	6.8872
#4	6.8758
#5	6.7493

Method: ISO_U Sample Name: s98t002827_a Operator: kjt
Comment: 40501 df
Run Time: 01/11/99 14:25 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.343	5.259	.1343	6.059	20.04
Stddev	.551	.564	.0075	.464	.38
%RSD	12.68	10.73	5.567	7.653	1.904
#1	3.936	5.612	.1264	5.612	20.00
#2	4.386	5.824	.1258	6.471	19.88
#3	4.574	4.646	.1391	5.517	19.77
#4	3.709	4.654	.1403	6.472	19.85
#5	5.112	5.556	.1397	6.223	20.71

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.8652
Stddev	.0740
%RSD	1.0781

#1	6.8599
#2	6.9539
#3	6.8872
#4	6.8758
#5	6.7493

Method: ISO_U Sample Name: ccv

Operator: kjt

Comment: iso-u lcs 20 ppb

Run Time: 01/11/99 14:29 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.853	5.521	38.31	5.206	4426.
Stddev	.512	.353	2.60	.744	133.
%RSD	10.54	6.387	6.783	14.28	2.998
#1	4.901	5.185	37.93	3.978	4416.
#2	4.468	5.117	39.85	5.045	4470.
#3	5.137	5.651	40.07	5.651	4481.
#4	4.240	5.921	33.92	5.556	4558.
#5	5.516	5.731	39.76	5.803	4207.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9218
Stddev	.0947
%RSD	1.3680

#1	7.0391
#2	6.9379
#3	6.8131
#4	6.8397
#5	6.9791

Method: ISO_U Sample Name: ccv

Operator: kjt

Comment: iso-u lcs 20 ppb

Run Time: 01/11/99 14:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.853	5.521	.1342	5.206	20.34
Stddev	.512	.353	.0109	.744	.61
%RSD	10.54	6.387	8.097	14.28	3.002
#1	4.901	5.185	.1326	3.978	20.29
#2	4.468	5.117	.1406	5.045	20.54
#3	5.137	5.651	.1415	5.651	20.59
#4	4.240	5.921	.1158	5.556	20.94
#5	5.516	5.731	.1402	5.803	19.33

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.9218
Stddev	.0947
%RSD	1.3680

#1	7.0391
#2	6.9379
#3	6.8131
#4	6.8397
#5	6.9791

HNF-1663 REV. 0

Method: ISO_U

Sample Name: ccb

Operator: kjt

Comment: 2% hno3

Run Time: 01/11/99 14:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.274	4.339	4.812	4.517	4.985
Stddev	.492	.288	.561	.430	.640
%RSD	11.52	6.647	11.66	9.517	12.85

#1	3.689	4.107	4.385	4.873	5.569
#2	4.708	4.299	5.254	4.299	4.367
#3	3.788	4.719	4.121	4.320	4.320
#4	4.604	4.539	5.447	5.058	4.993
#5	4.581	4.034	4.854	4.034	3.674

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.4116
Stddev	.2069
%RSD	2.7912

#1	7.1828
#2	7.3273
#3	7.5233
#4	7.7110
#5	7.3136

Method: ISO_U Sample Name: ccb

Operator: kjt

Comment: 2% hno3

Run Time: 01/11/99 14:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.274	4.339	-.0059	4.517	-.0059
Stddev	.492	.288	.0023	.430	.0029
%RSD	11.52	6.647	39.97	9.517	49.84
#1	3.689	4.107	-.0077	4.873	-.0032
#2	4.708	4.299	-.0040	4.299	-.0088
#3	3.788	4.719	-.0088	4.320	-.0090
#4	4.604	4.539	-.0032	5.058	-.0059
#5	4.581	4.034	-.0057	4.034	-.0027

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.4116
Stddev	.2069
%RSD	2.7912

#1	7.1828
#2	7.3273
#3	7.5233
#4	7.7110
#5	7.3136

LABCORE Completed Worklist Report for Worklist# 27496

Analyst: bjg

Instrument: ICPMS1

Book#: 42857

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		MSPR-QC TC99	QC	0.0197	1.55e-02	78.680 % Recovery	
1 ICV	0		MSPR-QC SN117	QC	0.0077	6.30e-03	81.818 % Recovery	
1 ICV	0		MSPR-QC SN126	QC	n/a	<1.50e-4	% Recovery	
1 ICV	0		MSPR-QC CS133	QC	0.0200	2.11e-02	105.500 % Recovery	
1 ICV	0		MSPR-QC CS135	QC	n/a	3.90e-05	% Recovery	
1 ICV	0		MSPR-QC CS137	QC	n/a	6.28e-05	% Recovery	
2 ICB	0		MSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2 ICB	0		MSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2 ICB	0		MSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2 ICB	0		MSPR-QC CS133	QC	1	1.24e-05	1.240e-005 ug/mL	
2 ICB	0		MSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2 ICB	0		MSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 TC99-D1	LIQUID	N/A	4.880e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 SN117-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 SN126-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 CS133-D1	LIQUID	N/A	7.157e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 CS135-D1	LIQUID	N/A	2.045e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T002824	0 D	MSPR-D1 CS137-D1	LIQUID	N/A	1.867e+00	8.10e-002 ug/mL	
4 DUP	S98T002824	0 D	MSPR-D1 TC99-D1	LIQUID	4.88e+00	5.05e+00	3.424 RPD	
4 DUP	S98T002824	0 D	MSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0	RPD	
4 DUP	S98T002824	0 D	MSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0	RPD	
4 DUP	S98T002824	0 D	MSPR-D1 CS133-D1	LIQUID	7.16e+00	7.27e+00	1.525 RPD	
4 DUP	S98T002824	0 D	MSPR-D1 CS135-D1	LIQUID	2.04e+00	2.06e+00	0.976 RPD	
4 DUP	S98T002824	0 D	MSPR-D1 CS137-D1	LIQUID	1.87e+00	1.91e+00	2.116 RPD	
5 SAMPLE	S98T002825	0 D	MSPR-D1 TC99-D1	LIQUID	N/A	4.560e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T002825	0 D	MSPR-D1 SN117-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
5 SAMPLE	S98T002825	0 D	MSPR-D1 SN126-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
5 SAMPLE	S98T002825	0 D	MSPR-D1 CS133-D1	LIQUID	N/A	6.229e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T002825	0 D	MSPR-D1 CS135-D1	LIQUID	N/A	1.827e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T002825	0 D	MSPR-D1 CS137-D1	LIQUID	N/A	1.628e+00	8.10e-002 ug/mL	
6 DUP	S98T002825	0 D	MSPR-D1 TC99-D1	LIQUID	4.56e+00	4.46e+00	2.217 RPD	
6 DUP	S98T002825	0 D	MSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0	RPD	
6 DUP	S98T002825	0 D	MSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0	RPD	
6 DUP	S98T002825	0 D	MSPR-D1 CS133-D1	LIQUID	6.23e+00	6.14e+00	1.455 RPD	
6 DUP	S98T002825	0 D	MSPR-D1 CS135-D1	LIQUID	1.83e+00	1.77e+00	3.333 RPD	
6 DUP	S98T002825	0 D	MSPR-D1 CS137-D1	LIQUID	1.63e+00	1.58e+00	3.115 RPD	
7 SAMPLE	S98T002940	0 D	MSPR-D1 TC99-D1	LIQUID	N/A	4.212e+00	8.10e-002 ug/mL	
7 SAMPLE	S98T002940	0 D	MSPR-D1 SN117-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
7 SAMPLE	S98T002940	0 D	MSPR-D1 SN126-D1	LIQUID	N/A	<6.075e+00	6.075 ug/mL	
7 SAMPLE	S98T002940	0 D	MSPR-D1 CS133-D1	LIQUID	N/A	7.023e+00	8.10e-002 ug/mL	
7 SAMPLE	S98T002940	0 D	MSPR-D1 CS135-D1	LIQUID	N/A	2.211e+00	8.10e-002 ug/mL	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27496

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7	SAMPLE	S98T002940	0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	1.847e+00	8.10e-002 ug/mL
8	DUP	S98T002940	0 D	QMSPR-D1 TC99-D1	LIQUID	4.21e+00	4.16e+00	1.195 RPD
8	DUP	S98T002940	0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0	RPD
8	DUP	S98T002940	0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0	RPD
8	DUP	S98T002940	0 D	QMSPR-D1 CS133-D1	LIQUID	7.02e+00	6.83e+00	2.744 RPD
8	DUP	S98T002940	0 D	QMSPR-D1 CS135-D1	LIQUID	2.21e+00	2.11e+00	4.630 RPD
8	DUP	S98T002940	0 D	QMSPR-D1 CS137-D1	LIQUID	1.85e+00	1.75e+00	5.556 RPD
9	SPK-POST	S98T002940	0 D	QMSPR-D1 TC99-D1	LIQUID	0.0219	1.78e-02	81.279 % Recovery
9	SPK-POST	S98T002940	0 D	QMSPR-D1 SN117-D1	LIQUID	n/a	4.25e-03	% Recovery
9	SPK-POST	S98T002940	0 D	QMSPR-D1 SN126-D1	LIQUID	n/a	8.00e-07	% Recovery
9	SPK-POST	S98T002940	0 D	QMSPR-D1 CS133-D1	LIQUID	0.0201	2.11e-02	104.975 % Recovery
9	SPK-POST	S98T002940	0 D	QMSPR-D1 CS135-D1	LIQUID	n/a	-3.00e-7	% Recovery
9	SPK-POST	S98T002940	0 D	QMSPR-D1 CS137-D1	LIQUID	n/a	9.00e-07	% Recovery
10	CCV	0	QMSPR-QC TC99	QC	0.0197	1.61e-02	81.726 % Recovery	
10	CCV	0	QMSPR-QC SN117	QC	0.0077	6.36e-03	82.597 % Recovery	
10	CCV	0	QMSPR-QC SN126	QC	n/a	<1.50e-4	% Recovery	
10	CCV	0	QMSPR-QC CS133	QC	0.0200	2.14e-02	107.000 % Recovery	
10	CCV	0	QMSPR-QC CS135	QC	n/a	3.94e-05	% Recovery	
10	CCV	0	QMSPR-QC CS137	QC	n/a	6.73e-05	% Recovery	
11	CCB	0	QMSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
11	CCB	0	QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
11	CCB	0	QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
11	CCB	0	QMSPR-QC CS133	QC	1	3.50e-06	3.500e-006 ug/mL	
11	CCB	0	QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
11	CCB	0	QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	

Final page for worklist# 27496

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. Pang 1/25/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27496

Analyst: B. Collier Instrument: ICPMS1 Book# 42857, 980121-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC QC			
2 ICB			@MSPR-QC QC			
3 SAMPLE	S98T002824 0 D		@MSPR-D1 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
4 DUP	S98T002824 0 D		@MSPR-D1 LIQUID			
5 SAMPLE	S98T002825 0 D		@MSPR-D1 LIQUID		98000437 BY-105 (2)	
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
6 DUP	S98T002825 0 D		@MSPR-D1 LIQUID			
7 SAMPLE	S98T002940 0 D		@MSPR-D1 LIQUID		98000490 BY-105 (2)	
	Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1					
8 DUP	S98T002940 0 D		@MSPR-D1 LIQUID			
9 SPK-POST	S98T002940 0 D		@MSPR-D1 LIQUID			
10 CCV			@MSPR-QC QC			
11 CCB			@MSPR-QC QC			

Final page for worklist # 27496

Analyst Signature B. Collier Date 1-22-99

Analyst Signature _____ Date _____

S98T002824, 2825, 2840

1-10-1 - 2025 - 10-1 40501

POST spike = 1 out of 44857

Data Entry Comments:

SCANS = PL2824, PL 2825, PL2840

Upload & Validation by:
Sant M. Pang
1/25/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2361

Analysis Report

01/22/99 14:01:46

page 1

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/22/99 14:00 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	98330.
Stddev	1386.
%RSD	1.410

#1	100600.
#2	98670.
#3	97410.
#4	97690.
#5	98600.
#6	98340.
#7	98460.
#8	95290.
#9	99540.
#10	98720.

[Signature]
BY-105
598700 2824, 2825, 2840
.1-10-.025-10
File # PRO1224.74T
CIL # 27486

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2362 TO 2378

2362

Method: TC_SN_CS Sample Name: Blank

Operator:

Comment:

Run Time: 01/22/99 14:02 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.845	1.726	9.769	2.303	1.052
Stddev	.261	.271	1.241	.313	.140
%RSD	14.12	15.69	12.70	13.61	13.34

#1	1.982	1.586	10.75	1.982	.8424
#2	1.751	1.651	8.005	2.702	1.101
#3	1.734	2.180	11.15	2.527	1.041
#4	2.216	1.478	9.457	2.019	1.231
#5	1.539	1.738	9.483	2.284	1.043

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.121	7.559	13.01	72.12
Stddev	.302	.499	.96	1.97
%RSD	4.942	6.604	7.407	2.729

#1	6.244	7.284	13.97	70.46
#2	5.804	7.755	13.21	73.60
#3	6.491	7.532	12.09	73.83
#4	5.812	8.275	11.92	69.55
#5	6.256	6.951	13.85	73.13

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.079
Stddev	.057
%RSD	.56421

#1	10.090
#2	9.9932
#3	10.091
#4	10.151
#5	10.070

Analysis Report

01/22/99 14:09:28

page 1

Method: TC_SN_CS Sample Name: PRIV Std

Operator:

Comment:

Run Time: 01/22/99 14:07 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]
Line	99/pulse	100/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	12580.	12390.	7194.	7194.
Stddev	380.	441.	122.	122.
%RSD	3.019	3.562	1.703	1.703

#1	12630.	12460.	7052.	7052.
#2	13160.	13140.	7317.	7317.
#3	12630.	12130.	7195.	7195.
#4	12190.	12170.	7315.	7315.
#5	12300.	12080.	7093.	7093.

Elem	Te[125]	Cs[133]	Cs[135]	Cs[137]
Line	125/pulse	133/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1505.	81050.	5761.	10080.
Stddev	44.	1656.	83.	196.
%RSD	2.954	2.044	1.445	1.947

#1	1508.	80680.	5820.	10210.
#2	1578.	83790.	5815.	10270.
#3	1459.	81190.	5827.	10130.
#4	1482.	79520.	5691.	9784.
#5	1499.	80050.	5651.	9985.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	60680.
Stddev	1577.
%RSD	2.599

#1	61330.
#2	62320.
#3	61650.
#4	58510.
#5	59600.

Analysis Report

01/22/99 14:09:28

page 2

Sample Name: PRIV Std Run Time: 01/22/99 14:07

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.9825
Stddev	.1739
%RSD	1.9360

#1	9.0198
#2	8.7437
#3	8.9274
#4	9.2261
#5	8.9953

2365

Analysis Report

01/22/99 14:18:42

page 1

Method: TC_SN_CS Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/22/99 14:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0002	-.0005	.0006	.0003
Stddev	.0004	.0003	.0018	.0006	.0008
%RSD	299.0	109.0	351.8	94.25	244.3

#1	-.0004	.0001	-.0013	.0002	.0002
#2	.0003	.0004	.0021	.0012	.0015
#3	-.0001	.0002	-.0023	.0011	-.0008
#4	.0003	.0006	.0004	.0004	.0004
#5	.0006	-.0001	-.0015	.0000	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0158	-.0040	-.0069	-.0049
Stddev	.0036	.0005	.0013	.0003
%RSD	22.79	13.00	19.55	6.789

#1	.0130	-.0043	-.0068	-.0051
#2	.0219	-.0035	-.0051	-.0045
#3	.0140	-.0040	-.0086	-.0050
#4	.0140	-.0035	-.0077	-.0052
#5	.0160	-.0047	-.0062	-.0045

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.2668
Stddev	.9138
%RSD	9.8615

#1	10.028
#2	7.7210
#3	9.7189
#4	9.6700
#5	9.1963

2366

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/22/99 14:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0037	.0004	6.298	.0001	.0015
Stddev	.0005	.0003	.152	.0005	.0012
%RSD	14.01	76.55	2.413	434.4	83.21

#1	.0032	.0002	6.552	.0008	.0029
#2	.0040	.0005	6.169	.0000	.0013
#3	.0041	.0001	6.266	-.0004	.0017
#4	.0030	.0003	6.197	.0001	.0021
#5	.0040	.0008	6.306	.0001	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0188	-.0041	-.0092	-.0068
Stddev	.0021	.0005	.0010	.0001
%RSD	10.92	12.96	10.70	1.467

#1	.0217	-.0050	-.0075	-.0067
#2	.0186	-.0040	-.0095	-.0067
#3	.0191	-.0036	-.0094	-.0069
#4	.0159	-.0042	-.0101	-.0068
#5	.0187	-.0038	-.0097	-.0067

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.0557
Stddev	.2443
%RSD	2.6973

#1	8.7231
#2	9.0823
#3	9.2426
#4	9.3217
#5	8.9086

Analysis Report

01/22/99 14:28:29

page 1

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/22/99 14:26 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	15.50	.0082	1.292	-.0001	-.0006
Stddev	.25	.0035	.071	.0002	.0019
%RSD	1.624	42.75	5.493	120.1	294.3

#1	15.33	.0145	1.193	.0000	-.0008
#2	15.27	.0066	1.271	-.0003	-.0028
#3	15.56	.0072	1.373	-.0004	.0021
#4	15.46	.0062	1.276	-.0001	.0002
#5	15.91	.0067	1.347	.0000	-.0018

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.09	.0390	.0628	.0437
Stddev	.17	.0006	.0016	.0010
%RSD	.7838	1.452	2.562	2.354

#1	21.05	.0392	.0623	.0446
#2	21.17	.0389	.0643	.0432
#3	20.84	.0395	.0631	.0425
#4	21.29	.0395	.0641	.0449
#5	21.13	.0381	.0603	.0432

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.6915
Stddev	.0452
%RSD	.46679

#1	9.7134
#2	9.6856
#3	9.7537
#4	9.6331
#5	9.6718

2368

Method: TC_SN_CS Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/22/99 14:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0000	-.0018	-.0001	.0000
Stddev	.0004	.0004	.0006	.0001	.0008
%RSD	54.70	2303.	32.68	153.5	4458.

#1	.0001	-.0003	-.0015	-.0002	-.0006
#2	.0011	.0003	-.0026	.0001	-.0007
#3	.0009	-.0005	-.0020	-.0001	.0006
#4	.0010	.0003	-.0011	.0000	.0010
#5	.0005	.0002	-.0017	-.0002	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0124	-.0040	-.0082	-.0052
Stddev	.0015	.0003	.0005	.0002
%RSD	12.00	6.195	6.416	3.413

#1	.0143	-.0041	-.0080	-.0051
#2	.0121	-.0044	-.0081	-.0055
#3	.0106	-.0038	-.0085	-.0052
#4	.0133	-.0040	-.0089	-.0050
#5	.0113	-.0040	-.0075	-.0052

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.006
Stddev	.569
%RSD	5.6848

#1	9.8353
#2	9.4885
#3	10.983
#4	9.8610
#5	9.8615

2369

Analysis Report

01/22/99 14:39:01

page 1

Method: TC_SN_CS Sample Name: s98t002824

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1205	.1578	.0334	.0010	-.0004
Stddev	.0049	.0062	.0009	.0006	.0010
%RSD	4.092	3.954	2.687	61.81	228.5
#1	.1152	.1515	.0327	.0017	-.0001
#2	.1218	.1528	.0331	.0000	.0011
#3	.1282	.1664	.0350	.0009	-.0013
#4	.1195	.1617	.0335	.0012	-.0014
#5	.1179	.1567	.0330	.0009	-.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1767	.0505	.0461	-.0073
Stddev	.0090	.0041	.0012	.0001
%RSD	5.075	8.157	2.618	.9536
#1	.1683	.0513	.0458	-.0072
#2	.1721	.0483	.0457	-.0072
#3	.1881	.0568	.0446	-.0073
#4	.1845	.0502	.0468	-.0073
#5	.1705	.0458	.0478	-.0074

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.3801
Stddev	.2419
%RSD	2.5794

#1	9.7422
#2	9.3767
#3	9.0703
#4	9.3016
#5	9.4098

2370

Analysis Report

01/22/99 14:41:17

page 1

Method: TC_SN_CS Sample Name: s98t002824_d

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1246	.1572	.0373	.0010	-.0010
Stddev	.0067	.0066	.0015	.0004	.0010
%RSD	5.392	4.201	3.943	42.10	103.8
#1	.1235	.1600	.0365	.0005	-.0019
#2	.1171	.1562	.0353	.0008	-.0020
#3	.1276	.1470	.0376	.0012	-.0008
#4	.1204	.1578	.0375	.0008	-.0006
#5	.1344	.1650	.0393	.0016	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1795	.0509	.0471	-.0075
Stddev	.0068	.0029	.0041	.0001
%RSD	3.807	5.730	8.710	1.193
#1	.1783	.0508	.0498	-.0076
#2	.1698	.0501	.0453	-.0074
#3	.1798	.0507	.0501	-.0074
#4	.1808	.0475	.0407	-.0075
#5	.1890	.0556	.0496	-.0074

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1569
Stddev	.3467
%RSD	3.7861

#1	9.1204
#2	9.7167
#3	9.0049
#4	9.1636
#5	8.7789

2371

Analysis Report

01/22/99 14:46:12

page 1

Method: TC_SN_CS Sample Name: s98t002825

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1126	.1413	.0308	.0009	-.0007
Stddev	.0041	.0059	.0021	.0002	.0014
%RSD	3.656	4.152	6.887	23.48	187.1
#1	.1187	.1517	.0305	.0010	.0014
#2	.1086	.1388	.0317	.0008	-.0003
#3	.1092	.1383	.0273	.0008	-.0012
#4	.1140	.1400	.0313	.0006	-.0023
#5	.1123	.1379	.0330	.0011	-.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1538	.0451	.0402	-.0077
Stddev	.0048	.0012	.0035	.0001
%RSD	3.150	2.565	8.825	1.054
#1	.1573	.0453	.0428	-.0075
#2	.1453	.0447	.0351	-.0078
#3	.1565	.0445	.0379	-.0076
#4	.1552	.0471	.0417	-.0077
#5	.1547	.0441	.0433	-.0076

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.7953
Stddev	.2873
%RSD	2.9328

#1	9.6689
#2	10.203
#3	9.9842
#4	9.5934
#5	9.5273

2372

Method: TC_SN_CS Sample Name: s98t002825_d

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:46 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1102	.1439	.0316	.0008	-.0004
Stddev	.0045	.0021	.0009	.0006	.0008
%RSD	4.123	1.456	2.787	74.62	184.4
#1	.1104	.1416	.0312	.0012	-.0010
#2	.1148	.1453	.0313	.0015	.0007
#3	.1075	.1467	.0323	-.0001	-.0011
#4	.1040	.1421	.0305	.0005	-.0010
#5	.1142	.1439	.0327	.0010	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1517	.0438	.0389	-.0076
Stddev	.0077	.0038	.0031	.0001
%RSD	5.057	8.653	7.854	.6720
#1	.1445	.0378	.0347	-.0076
#2	.1641	.0444	.0405	-.0076
#3	.1532	.0460	.0374	-.0075
#4	.1470	.0431	.0427	-.0076
#5	.1495	.0477	.0393	-.0076

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8446
Stddev	.2429
%RSD	2.4672

#1	10.164
#2	9.6310
#3	9.7838
#4	10.026
#5	9.6177

Analysis Report

01/22/99 14:53:00

page 1

Method: TC_SN_CS Sample Name: s98t002940

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:50 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1040	.1320	.0289	.0012	-.0003
Stddev	.0050	.0028	.0016	.0006	.0011
%RSD	4.767	2.089	5.433	48.56	318.2

#1	.1048	.1349	.0286	.0007	.0001
#2	.0989	.1285	.0309	.0015	-.0011
#3	.1065	.1346	.0298	.0007	-.0009
#4	.0994	.1303	.0286	.0011	-.0013
#5	.1106	.1314	.0267	.0022	.0014

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1734	.0546	.0456	-.0072
Stddev	.0052	.0031	.0021	.0001
%RSD	2.990	5.705	4.677	1.239

#1	.1761	.0570	.0467	-.0071
#2	.1653	.0530	.0485	-.0072
#3	.1785	.0528	.0444	-.0073
#4	.1714	.0515	.0453	-.0073
#5	.1757	.0588	.0430	-.0073

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4313
Stddev	.1691
%RSD	1.7927

#1	9.3224
#2	9.6777
#3	9.3444
#4	9.5343
#5	9.2777

2374

Method: TC_SN_CS Sample Name: s98t002940_d

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1027	.1287	.0287	.0009	-.0010
Stddev	.0069	.0086	.0017	.0003	.0007
%RSD	6.764	6.718	5.882	34.71	68.00

#1	.1074	.1352	.0305	.0005	-.0014
#2	.1087	.1341	.0297	.0014	-.0019
#3	.1048	.1358	.0294	.0009	-.0011
#4	.1011	.1191	.0267	.0009	-.0002
#5	.0914	.1195	.0270	.0008	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1687	.0521	.0432	-.0075
Stddev	.0096	.0022	.0056	.0001
%RSD	5.697	4.239	13.03	1.636

#1	.1758	.0552	.0484	-.0074
#2	.1719	.0523	.0477	-.0075
#3	.1739	.0529	.0437	-.0075
#4	.1700	.0510	.0420	-.0076
#5	.1520	.0493	.0343	-.0077

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4612
Stddev	.3377
%RSD	3.5693

#1	9.2160
#2	9.3826
#3	9.2741
#4	9.3820
#5	10.052

2075

Analysis Report

01/22/99 15:00:57

page 1

Method: TC_SN_CS Sample Name: s98t002940_a

Operator: bjg

Comment: 40501 df

Run Time: 01/22/99 14:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.88	.1320	4.250	.0008	-.0010
Stddev	.68	.0046	.138	.0004	.0009
%RSD	3.777	3.488	3.257	48.15	90.86
#1	18.39	.1311	4.248	.0013	-.0012
#2	16.83	.1260	4.100	.0007	.0004
#3	17.97	.1320	4.135	.0004	-.0014
#4	17.68	.1322	4.328	.0012	-.0019
#5	18.53	.1389	4.437	.0005	-.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.28	.0543	.0465	-.0065
Stddev	1.10	.0036	.0025	.0001
%RSD	5.176	6.656	5.283	1.701
#1	21.22	.0525	.0439	-.0066
#2	19.94	.0540	.0485	-.0065
#3	20.47	.0495	.0441	-.0065
#4	22.30	.0576	.0470	-.0065
#5	22.45	.0582	.0491	-.0063

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.3510
Stddev	.1949
%RSD	2.0841

#1	9.4323
#2	9.6032
#3	9.3548
#4	9.2938
#5	9.0710

2376

Analysis Report

01/22/99 15:09:41

page 1

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/22/99 15:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0085	-.0002	6.359	.0000	.0009
Stddev	.0009	.0003	.132	.000	.0010
%RSD	10.98	115.8	2.083	8226.	107.4
#1	.0089	-.0001	6.468	-.0001	.0017
#2	.0094	-.0004	6.382	.0006	.0007
#3	.0090	.0001	6.306	-.0003	.0009
#4	.0071	-.0006	6.479	.0000	-.0006
#5	.0079	-.0001	6.158	-.0003	.0018

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0048	-.0049	-.0105	-.0070
Stddev	.0015	.0004	.0005	.0001
%RSD	31.02	8.488	4.520	1.537
#1	.0072	-.0047	-.0109	-.0071
#2	.0047	-.0043	-.0100	-.0070
#3	.0036	-.0048	-.0100	-.0072
#4	.0048	-.0052	-.0108	-.0069
#5	.0036	-.0054	-.0109	-.0069

Int. Sti.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1761
Stddev	.0730
%RSD	.79580

#1	9.1532
#2	9.2292
#3	9.2651
#4	9.1549
#5	9.0780

Analysis Report

01/22/99 15:13:28

page 1

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/22/99 15:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.09	.0075	1.437	.0002	-.0005
Stddev	.22	.0013	.060	.0002	.0012
%RSD	1.373	17.36	4.207	103.2	231.8

#1	16.03	.0086	1.464	-.0001	-.0019
#2	16.40	.0073	1.392	.0003	.0000
#3	16.22	.0063	1.524	.0005	.0007
#4	15.88	.0091	1.433	.0001	.0002
#5	15.90	.0062	1.371	.0002	-.0016

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.36	.0394	.0673	.0457
Stddev	.22	.0019	.0030	.0024
%RSD	1.048	4.721	4.426	5.329

#1	21.61	.0412	.0674	.0443
#2	21.53	.0408	.0690	.0438
#3	21.40	.0379	.0669	.0460
#4	21.08	.0400	.0627	.0448
#5	21.20	.0369	.0706	.0499

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1031
Stddev	.0592
%RSD	.65085

#1	9.1230
#2	9.0055
#3	9.1126
#4	9.1081
#5	9.1663

2378

Analysis Report

01/22/99 15:21:18

page 1

Method: TC_SN_CS Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/22/99 15:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	-.0003	-.0020	-.0001	.0004
Stddev	.0003	.0002	.0010	.0004	.0008
%RSD	79.66	92.51	52.35	347.0	189.7

#1	.0004	.0002	-.0031	.0003	-.0005
#2	.0009	-.0004	-.0003	-.0008	.0006
#3	.0001	-.0004	-.0025	-.0002	.0009
#4	.0001	-.0003	-.0021	-.0002	.0015
#5	.0007	-.0004	-.0021	.0003	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0035	-.0042	-.0079	-.0050
Stddev	.0006	.0003	.0007	.0001
%RSD	18.47	8.092	8.577	2.013

#1	.0042	-.0041	-.0073	-.0049
#2	.0038	-.0043	-.0072	-.0050
#3	.0027	-.0048	-.0086	-.0049
#4	.0038	-.0039	-.0078	-.0048
#5	.0028	-.0040	-.0086	-.0051

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.2783
Stddev	.4479
%RSD	4.8276

#1	9.0654
#2	8.8769
#3	9.2890
#4	9.1253
#5	10.035

2379

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 8 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 2380 THRU 2707)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INDUCTIVELY COUPLED PLASMA/MASS SPECTROMETER ANALYSIS, ICP/MS WORKLIST 27497 THRU 27706		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 27497

Analyst: bjg

Instrument: ICPMS1

Book#: 27497

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSU-QC U235	QC	0.000714	7.18e-04	100.560 % Recovery	
1	ICV	0	EMSU-QC U238	QC	0.100	1.01e-01	101.000 % Recovery	
2	ICB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	SAMPLE	S98T002824 0 D	EMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002824 0 D	EMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002824 0 D	EMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002824 0 D	EMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
3	SAMPLE	S98T002824 0 D	EMSU-D1 U238-D1	LIQUID	N/A	7.695e-01	0.486 ug/mL	
4	DUP	S98T002824 0 D	EMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002824 0 D	EMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002824 0 D	EMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002824 0 D	EMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
4	DUP	S98T002824 0 D	EMSU-D1 U238-D1	LIQUID	7.69e-01	6.84e-01	11.700 RPD	
5	SAMPLE	S98T002825 0 D	EMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002825 0 D	EMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002825 0 D	EMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002825 0 D	EMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
5	SAMPLE	S98T002825 0 D	EMSU-D1 U238-D1	LIQUID	N/A	1.381e+00	0.486 ug/mL	
6	DUP	S98T002825 0 D	EMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002825 0 D	EMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002825 0 D	EMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002825 0 D	EMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
6	DUP	S98T002825 0 D	EMSU-D1 U238-D1	LIQUID	1.38e+00	1.30e+00	5.970 RPD	
7	SAMPLE	S98T002940 0 D	EMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002940 0 D	EMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002940 0 D	EMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002940 0 D	EMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
7	SAMPLE	S98T002940 0 D	EMSU-D1 U238-D1	LIQUID	N/A	2.531e+00	0.486 ug/mL	
8	DUP	S98T002940 0 D	EMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002940 0 D	EMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002940 0 D	EMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002940 0 D	EMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
8	DUP	S98T002940 0 D	EMSU-D1 U238-D1	LIQUID	2.53e+00	2.61e+00	3.113 RPD	
9	SPK-POST	S98T002940 0 D	EMSU-D1 U235-D1	LIQUID	0.0001422	1.43e-04	100.563 % Recovery	
9	SPK-POST	S98T002940 0 D	EMSU-D1 U238-D1	LIQUID	0.020	1.97e-02	98.500 % Recovery	
10	CCV	0	EMSU-QC U235	QC	0.0001422	1.40e-04	98.453 % Recovery	
10	CCV	0	EMSU-QC U238	QC	0.020	2.00e-02	100.000 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

01/26/99 07:06

Page: 1

A-0004-1

LABCORE Data Entry Template for Worklist# 27497

Analyst: BIG Instrument: ICPMS1 Book# 90855Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 SAMPLE	S98T002824	0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
4 DUP	S98T002824	0 D	@MSU-D1	LIQUID		
5 SAMPLE	S98T002825	0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
6 DUP	S98T002825	0 D	@MSU-D1	LIQUID		
7 SAMPLE	S98T002940	0 D	@MSU-D1	LIQUID	98000490	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
8 DUP	S98T002940	0 D	@MSU-D1	LIQUID		
9 SPK-POST	S98T002940	0 D	@MSU-D1	LIQUID		
10 CCV			@MSU-QC	QC		
11 CCB			@MSU-QC	QC		

Final page for worklist # 27497

Signature 2-16-99
Analyst Signature Date

Analyst Signature Date

S98T002824
 2824-D1
 2825
 2825-D
 2940
 2940-D
 2940-A

-1-10-1025-10

DF
 90501

Data Entry Comments:

Validated by:

POST Spike = total of 10 pp U STD

SCANS = 982824, 982825, 982940

2/18/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 02/16/99 13:04 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	28840.
Stddev	256.
%RSD	.8861

#1	28710.
#2	29060.
#3	29090.
#4	28450.
#5	29180.
#6	28560.
#7	28990.
#8	29020.
#9	28710.
#10	28620.

[Signature]
2-16-99

BY-105

11-10-05-10

ISO-U

File# 00216C.TM

W.C.F 27497

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2383 TO 241.

Method: ISO_U

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/16/99 13:06 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.413	4.443	4.696	4.905	5.433
Stddev	.423	.616	.628	.705	1.349
%RSD	9.589	13.86	13.37	14.38	24.83

#1	4.075	3.762	4.493	5.747	3.762
#2	4.693	4.805	5.811	5.476	6.817
#3	3.967	4.193	4.420	4.873	6.347
#4	4.979	5.318	4.300	4.187	5.997
#5	4.349	4.136	4.455	4.243	4.243

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.5608
Stddev	.1758
%RSD	3.8551

#1	4.7850
#2	4.4743
#3	4.4117
#4	4.4188
#5	4.7141

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	527.4941	4.4125	1.0000000	02/16/99 13:07:51

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.4125
100ppb U	0.711	0.711	0	0	379.46

U [233]
0
(S) IR⁴²⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	527.4513	4.4430	1.0000000	02/16/99 13:07:51

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.443
100ppb U	0.711	0.711	0	0	379.46

U [234]
0
(S) IR⁴²⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	527.0958	4.6957	1.0000000	02/16/99 13:07:51

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.6957
100ppb U	0.711	0.711	0	0	379.46

U [235]
0
(S) IR⁴²⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	526.8014	4.9051	1.0000000	02/16/99 13:07:51

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.9051
100ppb U	0.711	0.711	0	0	379.46

U [236]
0
(S) IR⁴²⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	511.3121	5.4329	1.0000000	02/16/99 13:07:51

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	5.4329
100ppb U	99.3	99.3	0	0	50779

U [238]

0

55860TIO

(S)IR

Concentration

Method: ISO_U

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/16/99 13:08 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	360.2	360.2	360.2	360.2	47140.
Stddev	9.7	9.7	9.7	9.7	448.
%RSD	2.690	2.690	2.690	2.690	.9501
#1	371.0	371.0	371.0	371.0	47730.
#2	362.3	362.3	362.3	362.3	47430.
#3	367.2	367.2	367.2	367.2	47010.
#4	348.5	348.5	348.5	348.5	47000.
#5	352.0	352.0	352.0	352.0	46560.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.5652
Stddev	.0604
%RSD	1.3224

#1	4.4661
#2	4.5596
#3	4.5778
#4	4.5986
#5	4.6240

HNF-1663 REV.0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	500.3976	4.4125	1.0000000	02/16/99 13:10:04

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4125
100ppb U	0.711	0.711	0	0	360.2

U [233]

0

(S) IR⁴⁰⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	500.3548	4.4430	1.0000000	02/16/99 13:10:04

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.443
100ppb U	0.711	0.711	0	0	360.2

U [234]

0

(S) IR⁴⁰⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	499.9993	4.6957	1.0000000	02/16/99 13:10:04

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.6957
100ppb U	0.711	0.711	0	0	360.2

U [235]

0

(S) IR⁴⁰⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	499.7049	4.9051	1.0000000	02/16/99 13:10:04

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.9051
100ppb U	0.711	0.711	0	0	360.2

U [236]

0

(S) IR⁴⁰⁰⁰¹⁰

Concentration

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U [238]	474.7121	5.4329	1.0000000	02/16/99 13:10:04

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	5.4329
100ppb U	99.3	99.3	0	0	47144

U [238]

0

(S) IR 51860 WU

Concentration

HNF-1663 REV. 0

Method: ISO_U Sample Name: blank

Operator: bjg

Comment: background

Run Time: 02/16/99 13:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	.0006	.0008	.0008	.0007
Stddev	.0021	.0014	.0012	.0011	.0009
%RSD	1113.	226.2	139.6	148.3	144.1

#1	.0009	.0011	-.0007	.0010	.0011
#2	.0025	.0016	.0004	.0022	.0017
#3	.0012	-.0019	.0004	.0000	-.0009
#4	-.0010	.0015	.0019	.0013	.0007
#5	-.0028	.0009	.0021	-.0007	.0006

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.5969
Stddev	.0205
%RSD	.44632

#1	4.6091
#2	4.5756
#3	4.5738
#4	4.6091
#5	4.6170

Analysis Report

02/16/99 13:25:00

HNF-1663 REV. 0

page 1

Method: ISO_U Sample Name: icv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/16/99 13:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0012	.0076	.7182	.0107	101.0
Stddev	.0014	.0017	.0174	.0017	.7
%RSD	115.8	22.53	2.428	16.19	.6776
#1	.0025	.0101	.7130	.0110	102.0
#2	.0004	.0073	.7429	.0097	101.1
#3	-.0001	.0054	.6984	.0089	100.1
#4	.0029	.0070	.7087	.0134	101.0
#5	.0004	.0079	.7282	.0103	101.0

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.5739
Stddev	.0347
%RSD	.75793

#1	4.5223
#2	4.5623
#3	4.6028
#4	4.6082
#5	4.5740

Method: ISO_U Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 02/16/99 13:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	.0007	.0006	.0003	-.0019
Stddev	.0006	.0015	.0017	.0009	.0019
%RSD	258.3	209.1	272.9	312.9	100.9

#1	.0008	.0001	-.0009	-.0006	-.0022
#2	.0006	-.0012	.0024	-.0004	-.0043
#3	.0003	.0004	.0008	.0004	-.0007
#4	-.0007	.0028	.0021	.0016	.0006
#5	.0001	.0016	-.0013	.0004	-.0027

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.5771
Stddev	.0298
%RSD	.65111

#1	4.5742
#2	4.5673
#3	4.6199
#4	4.5380
#5	4.5863

HNF-1663 REV. 0

Method: ISO_U Sample Name: s98t002824 Operator: bjg
 Comment: 40501 df
 Run Time: 02/16/99 13:31 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0013	.0007	.0021	-.0008	.0190
Stddev	.0011	.0020	.0008	.0014	.0039
%RSD	88.21	272.5	36.51	165.2	20.49

#1	.0026	-.0012	.0014	-.0002	.0243
#2	.0013	.0003	.0014	.0003	.0177
#3	.0016	.0040	.0026	-.0027	.0204
#4	-.0005	.0010	.0032	.0003	.0136
#5	.0014	-.0004	.0020	-.0020	.0191

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4649
Stddev	.0206
%RSD	.46107

#1	4.4525
#2	4.4409
#3	4.4925
#4	4.4599
#5	4.4784

Method: ISO_U

Sample Name: s98t002824_d

Operator: bjg

Comment: 40501 df

Run Time: 02/16/99 13:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	-.0007	-.0018	-.0012	.0169
Stddev	.0014	.0012	.0017	.0010	.0014
%RSD	2183.	163.6	92.74	82.36	8.425
#1	-.0008	-.0027	-.0009	-.0009	.0159
#2	.0007	.0000	-.0039	-.0028	.0156
#3	.0019	.0003	-.0024	-.0002	.0169
#4	-.0006	-.0004	-.0025	-.0007	.0167
#5	-.0016	-.0008	.0005	-.0017	.0192

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4270
Stddev	.0515
%RSD	1.1642

#1	4.3565
#2	4.3974
#3	4.4651
#4	4.4845
#5	4.4313

Method: ISO_U Sample Name: s98t002825

Operator: bjg

Comment: 40501 df

Run Time: 02/16/99 13:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	-.0002	-.0014	-.0016	.0341
Stddev	.0009	.0017	.0019	.0021	.0006
%RSD	180.0	720.1	139.6	127.7	1.774
#1	-.0008	-.0025	-.0023	-.0016	.0350
#2	-.0018	-.0005	.0012	-.0012	.0343
#3	-.0002	.0004	-.0016	.0015	.0338
#4	-.0003	.0021	-.0039	-.0027	.0334
#5	.0006	-.0008	-.0004	-.0042	.0341

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4462
Stddev	.0688
%RSD	1.5467

#1	4.5134
#2	4.4200
#3	4.5118
#4	4.3498
#5	4.4358

Method: ISO_U   Sample Name: s98t002825_d   Operator: bjg
Comment: 40501 df
Run Time: 02/16/99 13:40 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	.0000	-.0005	-.0017	.0322
Stddev	.0009	.001	.0020	.0009	.0035
%RSD	147.8	4872.	369.6	52.03	10.81
#1	-.0021	.0002	-.0024	-.0007	.0290
#2	-.0002	-.0009	-.0001	-.0014	.0289
#3	.0003	-.0002	-.0014	-.0030	.0362
#4	-.0004	-.0014	.0027	-.0011	.0316
#5	-.0007	.0022	-.0015	-.0021	.0354

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3793
Stddev	.0477
%RSD	1.0882

#1	4.3019
#2	4.4101
#3	4.3744
#4	4.3849
#5	4.4250

Method: ISO_U Sample Name: s98t002940

Operator: bjg

Comment: 40501 df

Run Time: 02/16/99 13:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0010	-.0010	-.0010	-.0024	.0625
Stddev	.0011	.0005	.0011	.0014	.0035
%RSD	113.6	53.37	110.7	57.98	5.579
#1	-.0013	-.0011	-.0023	-.0023	.0652
#2	-.0012	-.0001	-.0020	-.0015	.0669
#3	-.0018	-.0015	-.0006	-.0035	.0597
#4	-.0017	-.0013	.0005	-.0006	.0588
#5	.0010	-.0009	-.0007	-.0039	.0620

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4009
Stddev	.0425
%RSD	.96605

#1	4.3822
#2	4.4481
#3	4.4408
#4	4.3477
#5	4.3858

Method: ISO_U

Sample Name: s98t002940_d

Operator: bjg

Comment: 40501 df

Run Time: 02/16/99 13:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0012	.0003	-.0015	-.0016	.0645
Stddev	.0015	.0009	.0009	.0017	.0046
%RSD	125.3	275.8	59.85	104.5	7.102
#1	-.0029	.0004	-.0001	-.0010	.0611
#2	-.0016	-.0001	-.0020	-.0003	.0595
#3	.0011	.0010	-.0020	-.0037	.0633
#4	-.0006	-.0009	-.0022	.0001	.0699
#5	-.0018	.0013	-.0010	-.0030	.0686

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.3802
Stddev	.1250
%RSD	2.8541

#1	4.4162
#2	4.4390
#3	4.4530
#4	4.1578
#5	4.4350

Method: ISO_U   Sample Name: s98t002940_a   Operator: bjg
Comment: 40501 df
Run Time: 02/16/99 13:47 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	.0005	.1432	.0018	19.75
Stddev	.0016	.0021	.0082	.0013	.31
%RSD	989.7	389.4	5.704	69.13	1.550

#1	.0027	.0006	.1555	.0012	20.14
#2	-.0012	-.0008	.1451	.0034	19.38
#3	-.0012	.0012	.1359	.0003	19.61
#4	.0004	-.0019	.1440	.0015	19.97
#5	.0001	.0036	.1355	.0027	19.64

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4532
Stddev	.0202
%RSD	.45363

#1	4.4406
#2	4.4527
#3	4.4727
#4	4.4268
#5	4.4731

Method: ISO_U Sample Name: ccv

Operator: bjg

Comment: 20 ppb u lcs

Run Time: 02/16/99 13:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0011	.1400	.0019	20.03
Stddev	.0008	.0015	.0028	.0008	.47
%RSD	325.4	141.9	1.984	42.88	2.351
#1	-.0008	.0016	.1391	.0018	19.96
#2	.0002	.0021	.1445	.0023	20.04
#3	.0012	-.0002	.1384	.0020	19.50
#4	-.0003	.0028	.1408	.0007	20.78
#5	.0010	-.0009	.1374	.0029	19.85

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.4214
Stddev	.1240
%RSD	2.8046

#1	4.4932
#2	4.4513
#3	4.4799
#4	4.2013
#5	4.4810

Method: ISO_U Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/16/99 13:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0003	-.0015	-.0008	-.0020
Stddev	.0016	.0018	.0006	.0019	.0010
%RSD	171.7	568.2	40.31	236.7	49.39
#1	-.0023	-.0012	-.0022	-.0019	-.0014
#2	.0012	.0028	-.0017	-.0030	-.0031
#3	-.0025	-.0005	-.0014	.0021	-.0006
#4	-.0007	-.0012	-.0006	-.0010	-.0021
#5	-.0002	-.0014	-.0015	-.0003	-.0026

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	4.2938
Stddev	.0139
%RSD	.32274

#1	4.3112
#2	4.2805
#3	4.2792
#4	4.3022
#5	4.2960

LABCORE Completed Worklist Report for Worklist# 27498

Analyst: kjt Instrument: ICPMS1

Book#: 48B57

Method: LA-506-101 Rev/Mod A-0

49B57

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	QMSPR-QC TC99	QC	0.0205	1.75e-02	85.366 % Recovery	
1 ICV	0	QMSPR-QC SN117	QC	0.01925	2.03e-02	105.455 % Recovery	
1 ICV	0	QMSPR-QC CS133	QC	0.0200	1.78e-02	89.000 % Recovery	
2 ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
2 ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
2 ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
2 ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
2 ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
2 ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	5.403e+00	8.10e-002	ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	6.010e+00	8.10e-002	ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.122e+00	8.10e-002	ug/mL
3 SAMPLE	S98T002826 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.074e+00	8.10e-002	ug/mL
4 DUP	S98T002826 0 D	QMSPR-D1 TC99-D1	LIQUID	5.40e+00	5.40e+00	0.000	RPD
4 DUP	S98T002826 0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
4 DUP	S98T002826 0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
4 DUP	S98T002826 0 D	QMSPR-D1 CS133-D1	LIQUID	6.01e+00	6.00e+00	0.167	RPD
4 DUP	S98T002826 0 D	QMSPR-D1 CS135-D1	LIQUID	2.12e+00	2.11e+00	0.473	RPD
4 DUP	S98T002826 0 D	QMSPR-D1 CS137-D1	LIQUID	2.07e+00	2.11e+00	1.914	RPD
5 SAMPLE	S98T002828 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	5.172e+00	8.10e-002	ug/mL
5 SAMPLE	S98T002828 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
5 SAMPLE	S98T002828 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
5 SAMPLE	S98T002828 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	6.355e+00	8.10e-002	ug/mL
5 SAMPLE	S98T002828 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.369e+00	8.10e-002	ug/mL
5 SAMPLE	S98T002828 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.268e+00	8.10e-002	ug/mL
6 DUP	S98T002828 0 D	QMSPR-D1 TC99-D1	LIQUID	5.17e+00	5.10e+00	1.363	RPD
6 DUP	S98T002828 0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
6 DUP	S98T002828 0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
6 DUP	S98T002828 0 D	QMSPR-D1 CS133-D1	LIQUID	6.36e+00	6.33e+00	0.473	RPD
6 DUP	S98T002828 0 D	QMSPR-D1 CS135-D1	LIQUID	2.37e+00	2.30e+00	2.998	RPD
6 DUP	S98T002828 0 D	QMSPR-D1 CS137-D1	LIQUID	2.27e+00	2.24e+00	1.330	RPD
7 SAMPLE	S98T002965 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	5.067e+00	8.10e-002	ug/mL
7 SAMPLE	S98T002965 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
7 SAMPLE	S98T002965 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	< 6.075e+00	6.075	ug/mL
7 SAMPLE	S98T002965 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	5.946e+00	8.10e-002	ug/mL
7 SAMPLE	S98T002965 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.126e+00	8.10e-002	ug/mL
7 SAMPLE	S98T002965 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.049e+00	8.10e-002	ug/mL
8 DUP	S98T002965 0 D	QMSPR-D1 TC99-D1	LIQUID	5.07e+00	5.07e+00	0.000	RPD
8 DUP	S98T002965 0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

2402

LABCORE Completed Worklist Report for Worklist# 27498

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
8 DUP	S98T002965	0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
8 DUP	S98T002965	0 D	QMSPR-D1 CS133-D1	LIQUID	5.95e+00	5.89e+00	1.014	RPD
8 DUP	S98T002965	0 D	QMSPR-D1 CS135-D1	LIQUID	2.13e+00	2.13e+00	0.000	RPD
8 DUP	S98T002965	0 D	QMSPR-D1 CS137-D1	LIQUID	2.05e+00	2.10e+00	2.410	RPD
9 SAMPLE	S98T003385	0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	4.868e+00	8.10e-002	ug/mL
9 SAMPLE	S98T003385	0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
9 SAMPLE	S98T003385	0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
9 SAMPLE	S98T003385	0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	5.921e+00	8.10e-002	ug/mL
9 SAMPLE	S98T003385	0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.090e+00	8.10e-002	ug/mL
9 SAMPLE	S98T003385	0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.021e+00	8.10e-002	ug/mL
10 DUP	S98T003385	0 D	QMSPR-D1 TC99-D1	LIQUID	4.87e+00	4.75e+00	2.495	RPD
10 DUP	S98T003385	0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
10 DUP	S98T003385	0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
10 DUP	S98T003385	0 D	QMSPR-D1 CS133-D1	LIQUID	5.92e+00	5.76e+00	2.740	RPD
10 DUP	S98T003385	0 D	QMSPR-D1 CS135-D1	LIQUID	2.09e+00	2.07e+00	0.962	RPD
10 DUP	S98T003385	0 D	QMSPR-D1 CS137-D1	LIQUID	2.02e+00	1.98e+00	2.000	RPD
11 SPK-POST	S98T003385	0 D	QMSPR-D1 TC99-D1	LIQUID	0.0219	1.83e-02	83.562	% Recovery
11 SPK-POST	S98T003385	0 D	QMSPR-D1 SN117-D1	LIQUID	0.0193	1.73e-02	89.637	% Recovery
11 SPK-POST	S98T003385	0 D	QMSPR-D1 CS133-D1	LIQUID	0.0201	1.80e-02	89.552	% Recovery
12 SAMPLE	S98T002391	0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	2.200e-05	2.20e-005	ug/mL
12 SAMPLE	S98T002391	0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	1.650e-03	1.65e-003	ug/mL
12 SAMPLE	S98T002391	0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	1.650e-03	1.65e-003	ug/mL
12 SAMPLE	S98T002391	0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	2.200e-05	2.20e-005	ug/mL
12 SAMPLE	S98T002391	0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.200e-05	2.20e-005	ug/mL
12 SAMPLE	S98T002391	0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	3.080e-05	2.20e-005	ug/mL
13 DUP	S98T002391	0 D	QMSPR-D1 TC99-D1	LIQUID	<2.20e-5	<2.20e-5		RPD
13 DUP	S98T002391	0 D	QMSPR-D1 SN117-D1	LIQUID	<1.65e-3	<1.65e-3		RPD
13 DUP	S98T002391	0 D	QMSPR-D1 SN126-D1	LIQUID	<1.65e-3	<1.65e-3		RPD
13 DUP	S98T002391	0 D	QMSPR-D1 CS133-D1	LIQUID	<2.20e-5	<2.20e-5		RPD
13 DUP	S98T002391	0 D	QMSPR-D1 CS135-D1	LIQUID	<2.20e-5	<2.20e-5		RPD
13 DUP	S98T002391	0 D	QMSPR-D1 CS137-D1	LIQUID	3.08e-05	2.86e-05	7.407	RPD
14 CCV		0	QMSPR-QC TC99	QC	0.0205	1.70e-02	82.927	% Recovery
14 CCV		0	QMSPR-QC SN117	QC	0.01925	2.01e-02	104.416	% Recovery
14 CCV		0	QMSPR-QC CS133	QC	0.0200	1.76e-02	88.000	% Recovery
15 CCB		0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
15 CCB		0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
15 CCB		0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
15 CCB		0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
15 CCB		0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
15 CCB		0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27498

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. Pang 03/08/99

LABCORE Data Entry Template for Worklist# 27498

Analyst: Rsc Instrument: ICPMS1 Book# 49857
 Method: LA-506-101 Rev/Mod A-O 49857

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV				@MSPR-QC	QC		
2	ICB				@MSPR-QC	QC		
3	SAMPLE	S98T002826	0	D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, SN117-D1, SN126-D1, TC99-D1						
4	DUP	S98T002826	0	D	@MSPR-D1	LIQUID		
5	SAMPLE	S98T002828	0	D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, SN117-D1, SN126-D1, TC99-D1						
6	DUP	S98T002828	0	D	@MSPR-D1	LIQUID		
7	SAMPLE	S98T002965	0	D	@MSPR-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, SN117-D1, SN126-D1, TC99-D1						
8	DUP	S98T002965	0	D	@MSPR-D1	LIQUID		
9	SAMPLE	S98T003385	0	D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, SN117-D1, SN126-D1, TC99-D1						
10	DUP	S98T003385	0	D	@MSPR-D1	LIQUID		
11	SPK-POST	S98T003385	0	D	@MSPR-D1	LIQUID		
12	SAMPLE	S98T002391	0	D	@MSPR-D1	LIQUID	98000436	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, SN117-D1, SN126-D1, TC99-D1						
13	DUP	S98T002391	0	D	@MSPR-D1	LIQUID		
14	CCV				@MSPR-QC	QC		

Data Entry Comments:

Uploaded & Validated by:
 Sam M. Pang
 03/08/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27498

S Type	Sample#	R A Test	Matrix	Group#	Project
15 CCB		@MSPR-QC QC			

Final page for worklist # 27498

K. S. Thomdike 03-05-99
Signature Date

_____
Signature Date

file pr 0305a.txt

S98T002826

.1-10-.025-10

40501 dF

2826-d

2828

2828-d

2965

2965-d

3385

3385-d

3385-a

3385-Q

S98T002391

1-10

11 dF

2391-d

↓

↓

Spike .1ml mspr spike 44857
.25ml of Sn 10ppm (44857) (990119-1) Sn

Scan pr 2826
pr 2828
pr 2965
pr 3385
pr 2391

Data Entry Comments:

HNF-1663 REV. 0

Method: MS_TUNE Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/05/99 13:00 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	186600.
Stddev	1185.
%RSD	.6350

#1	185100.
#2	187800.
#3	186700.
#4	185200.
#5	185900.
#6	185800.
#7	186200.
#8	188100.
#9	188300.
#10	186700.

K3 Phouclike 03-05-99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2404 TO 2428

Method: PRV Sample Name: blank

Operator: kjt

Comment: background

Run Time: 03/05/99 13:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0020	-.0034	.0025	.0001
Stddev	.0002	.0008	.0106	.0005	.0003
%RSD	46.36	39.73	313.4	18.37	560.9

#1	.0005	.0020	-.0009	.0023	.0005
#2	.0001	.0030	.0123	.0021	-.0002
#3	.0003	.0014	-.0165	.0021	.0002
#4	.0003	.0011	-.0083	.0028	-.0003
#5	.0004	.0024	-.0035	.0031	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0002	-.0001	-.0001	-.0003
Stddev	.0002	.0003	.0002	.0004
%RSD	90.16	611.6	161.4	141.4

#1	.0000	-.0003	-.0001	-.0005
#2	.0004	.0000	.0002	.0003
#3	.0003	.0000	-.0003	-.0008
#4	.0004	-.0004	-.0003	-.0003
#5	.0000	.0004	-.0003	-.0001

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.568
Stddev	.081
%RSD	.43669

#1	18.459
#2	18.534
#3	18.555
#4	18.624
#5	18.667

Method: PRV Sample Name: icv Operator: kjt
 Comment: msprv lcs
 Run Time: 03/05/99 13:06 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.49	.0012	.0026	1.328	.0000
Stddev	.16	.0006	.0091	.051	.000
%RSD	.9328	51.02	348.8	3.811	.702.1

#1	17.44	.0015	.0081	1.252	-.0002
#2	17.52	.0003	.0138	1.306	-.0003
#3	17.68	.0009	-.0012	1.338	.0000
#4	17.58	.0020	.0026	1.369	.0000
#5	17.25	.0014	-.0102	1.375	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	17.84	.0062	.0106	.0987
Stddev	.11	.0007	.0007	.0015
%RSD	.6117	11.20	6.138	1.490

#1	17.81	.0059	.0110	.0979
#2	17.84	.0062	.0103	.0983
#3	17.81	.0074	.0105	.1009
#4	18.02	.0055	.0114	.0971
#5	17.72	.0062	.0097	.0994

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.583
Stddev	.109
%RSD	.58769

#1	18.697
#2	18.574
#3	18.528
#4	18.434
#5	18.680

Method: PRV Sample Name: icv

Operator: kjt

Comment: 19.25 ppb sn

Run Time: 03/05/99 13:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.5311	.0004	.0058	20.30	.0000
Stddev	.6230	.0007	.0106	.26	.000
%RSD	117.3	162.9	180.6	1.293	292.9
#1	1.597	.0003	.0069	20.57	.0000
#2	.5641	.0005	-.0005	20.10	-.0002
#3	.2528	.0016	-.0081	20.50	-.0001
#4	.1422	-.0003	.0190	19.96	.0001
#5	.0992	.0001	.0119	20.35	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0089	.0259	.0415	.3804
Stddev	.0010	.0007	.0016	.0058
%RSD	11.56	2.818	3.809	1.524
#1	.0107	.0263	.0414	.3866
#2	.0086	.0253	.0425	.3807
#3	.0084	.0270	.0416	.3843
#4	.0085	.0252	.0389	.3792
#5	.0083	.0258	.0429	.3714

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.652
Stddev	.165
%RSD	.99042

#1	16.809
#2	16.799
#3	16.414
#4	16.660
#5	16.576

Method: PRV

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/05/99 13:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0012	-.0050	.0065	.0001
Stddev	.0002	.0011	.0097	.0006	.0003
%RSD	25.17	87.56	193.2	9.397	431.9
#1	.0008	.0008	-.0128	.0073	.0000
#2	.0007	.0018	.0116	.0058	.0000
#3	.0010	-.0004	-.0050	.0066	.0001
#4	.0013	.0023	-.0105	.0067	.0004
#5	.0007	.0016	-.0086	.0060	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0003	.0000	-.0003	-.0003
Stddev	.0003	.000	.0002	.0007
%RSD	86.69	639.0	83.05	287.7
#1	.0007	-.0002	-.0006	-.0003
#2	.0003	.0000	-.0002	.0000
#3	.0006	.0000	-.0002	-.0008
#4	.0001	-.0001	-.0004	-.0011
#5	.0000	.0002	.0000	.0008

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.640
Stddev	.110
%RSD	.59156

#1	18.461
#2	18.671
#3	18.621
#4	18.750
#5	18.696

Method: PRV Sample Name: s98t002826 Operator: kjt
 Comment: 40501 df
 Run Time: 03/05/99 13:33 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1334	.0823	-.0124	.0497	.0014
Stddev	.0028	.0032	.0097	.0019	.0002
%RSD	2.135	3.831	78.26	3.898	13.23
#1	.1354	.0799	-.0127	.0507	.0014
#2	.1290	.0823	.0034	.0474	.0015
#3	.1359	.0787	-.0121	.0525	.0013
#4	.1322	.0839	-.0191	.0490	.0016
#5	.1345	.0867	-.0216	.0492	.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1484	.0524	.0512	-.0053
Stddev	.0027	.0018	.0007	.0005
%RSD	1.853	3.361	1.348	9.009
#1	.1525	.0520	.0516	-.0058
#2	.1498	.0549	.0517	-.0057
#3	.1473	.0513	.0517	-.0052
#4	.1463	.0534	.0506	-.0046
#5	.1459	.0505	.0503	-.0051

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.217
Stddev	.287
%RSD	1.4958

#1	18.905
#2	19.015
#3	19.182
#4	19.348
#5	19.634

HNF-1663 REV. 0

Method: PRV

Sample Name: S98T002826_d

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1334	.0830	-.0117	.0491	.0011
Stddev	.0024	.0021	.0103	.0009	.0003
%RSD	1.832	2.575	88.08	1.744	26.69

#1	.1318	.0847	-.0059	.0504	.0011
#2	.1350	.0834	-.0205	.0492	.0008
#3	.1350	.0854	.0018	.0485	.0013
#4	.1355	.0811	-.0107	.0482	.0014
#5	.1299	.0806	-.0230	.0491	.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1481	.0520	.0520	-.0054
Stddev	.0019	.0017	.0014	.0008
%RSD	1.295	3.350	2.749	15.23

#1	.1493	.0529	.0534	-.0046
#2	.1459	.0536	.0517	-.0058
#3	.1506	.0503	.0498	-.0045
#4	.1468	.0532	.0522	-.0060
#5	.1478	.0499	.0531	-.0062

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.347
Stddev	.126
%RSD	.64888

#1	19.279
#2	19.270
#3	19.371
#4	19.259
#5	19.557

Method: PRV

Sample Name: s98t002828

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:39

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1277	.0820	-.0042	.0492	.0013
Stddev	.0056	.0024	.0060	.0014	.0005
%RSD	4.372	2.941	141.1	2.858	39.10

#1	.1376	.0791	-.0049	.0504	.0016
#2	.1258	.0808	.0030	.0509	.0009
#3	.1249	.0833	-.0020	.0491	.0019
#4	.1244	.0854	-.0135	.0483	.0013
#5	.1258	.0815	-.0038	.0475	.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1569	.0585	.0560	.0041
Stddev	.0061	.0029	.0017	.0010
%RSD	3.861	4.932	2.978	24.96

#1	.1641	.0631	.0586	.0058
#2	.1507	.0583	.0540	.0041
#3	.1626	.0557	.0557	.0031
#4	.1541	.0589	.0563	.0043
#5	.1529	.0564	.0555	.0034

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.265
Stddev	.604
%RSD	3.1336

#1	18.187
#2	19.481
#3	19.509
#4	19.583
#5	19.564

Method: PRV

Sample Name: s98t002828_d

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1259	.0763	-.0137	.0474	.0012
Stddev	.0021	.0022	.0061	.0015	.0001
%RSD	1.681	2.923	44.87	3.177	11.80

#1	.1254	.0789	-.0090	.0473	.0014
#2	.1285	.0745	-.0173	.0484	.0014
#3	.1249	.0785	-.0144	.0491	.0011
#4	.1232	.0745	-.0215	.0468	.0011
#5	.1276	.0750	-.0063	.0452	.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1562	.0568	.0554	.0033
Stddev	.0019	.0017	.0016	.0006
%RSD	1.189	3.070	2.949	18.26

#1	.1561	.0550	.0531	.0033
#2	.1567	.0563	.0543	.0037
#3	.1559	.0558	.0570	.0027
#4	.1536	.0571	.0560	.0028
#5	.1588	.0595	.0566	.0041

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.658
Stddev	.183
%RSD	.93195

#1	19.678
#2	19.766
#3	19.861
#4	19.605
#5	19.378

Analysis Report

HNF-1663 REV. 0

03/05/99 13:47:23

page 1

Method: PRV

Sample Name: s98t002965

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1251	.0778	-.0077	.0448	.0010
Stddev	.0035	.0027	.0079	.0007	.0003
%RSD	2.820	3.432	102.5	1.587	29.82

#1	.1211	.0740	-.0120	.0447	.0012
#2	.1230	.0761	-.0141	.0459	.0012
#3	.1287	.0803	.0033	.0451	.0005
#4	.1238	.0788	-.0020	.0443	.0010
#5	.1289	.0799	-.0137	.0441	.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1468	.0525	.0506	-.0062
Stddev	.0031	.0013	.0009	.0004
%RSD	2.110	2.457	1.770	6.690

#1	.1424	.0518	.0501	-.0064
#2	.1503	.0515	.0506	-.0057
#3	.1491	.0522	.0501	-.0067
#4	.1455	.0523	.0501	-.0064
#5	.1469	.0547	.0522	-.0058

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.836
Stddev	.171
%RSD	.86184

#1	20.086
#2	19.758
#3	19.765
#4	19.925
#5	19.648

Method: PRV Sample Name: s98t002965_d Operator: kjt
 Comment: 40501 df
 Run Time: 03/05/99 13:48 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1253	.0771	-.0204	.0456	.0010
Stddev	.0026	.0013	.0109	.0005	.0002
%RSD	2.086	1.641	53.44	1.028	23.04
#1	.1280	.0764	-.0140	.0452	.0012
#2	.1232	.0785	-.0331	.0455	.0009
#3	.1219	.0780	-.0068	.0453	.0007
#4	.1268	.0774	-.0184	.0457	.0013
#5	.1268	.0754	-.0297	.0464	.0011

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1454	.0525	.0519	-.0061
Stddev	.0027	.0014	.0012	.0009
%RSD	1.836	2.637	2.265	14.87
#1	.1426	.0506	.0509	-.0070
#2	.1445	.0536	.0533	-.0061
#3	.1436	.0528	.0504	-.0062
#4	.1492	.0539	.0522	-.0046
#5	.1471	.0518	.0525	-.0066

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.759
Stddev	.121
%RSD	.61173

#1	19.750
#2	19.782
#3	19.948
#4	19.635
#5	19.679

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t003385

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1202	.0735	-.0077	.0425	.0011
Stddev	.0059	.0042	.0061	.0031	.0002
%RSD	4.921	5.759	78.81	7.259	22.52

#1	.1306	.0800	-.0061	.0480	.0015
#2	.1189	.0689	-.0108	.0414	.0012
#3	.1165	.0707	-.0063	.0413	.0009
#4	.1184	.0740	.0005	.0419	.0009
#5	.1165	.0739	-.0159	.0402	.0010

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1462	.0516	.0499	-.0080
Stddev	.0074	.0028	.0023	.0002
%RSD	5.069	5.354	4.574	3.004

#1	.1590	.0555	-.0539	-.0079
#2	.1429	.0509	.0493	-.0081
#3	.1413	.0495	.0498	-.0079
#4	.1465	.0531	.0484	-.0084
#5	.1415	.0487	.0483	-.0078

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.427
Stddev	.761
%RSD	3.9165

#1	18.070
#2	19.745
#3	19.806
#4	19.685
#5	19.830

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t003385_d

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1172	.0758	-.0087	.0421	.0010
Stddev	.0022	.0011	.0152	.0012	.0004
%RSD	1.883	1.507	173.9	2.802	35.70

#1	.1199	.0766	-.0121	.0433	.0011
#2	.1166	.0755	-.0140	.0405	.0008
#3	.1178	.0739	.0129	.0415	.0014
#4	.1178	.0762	-.0281	.0420	.0015
#5	.1139	.0766	-.0024	.0431	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1423	.0511	.0488	-.0079
Stddev	.0019	.0018	.0016	.0006
%RSD	1.339	3.553	3.325	7.476

#1	.1442	.0499	.0498	-.0081
#2	.1419	.0532	.0507	-.0075
#3	.1399	.0496	.0468	-.0074
#4	.1442	.0530	.0490	-.0076
#5	.1412	.0500	.0475	-.0088

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.765
Stddev	.150
%RSD	.76146

#1	19.648
#2	19.728
#3	19.747
#4	19.676
#5	20.024

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t003385_a

Operator: kjt

Comment: 40501 df

Run Time: 03/05/99 13:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.41	.0726	-.0124	3.452	.0012
Stddev	.09	.0026	.0134	.023	.0001
%RSD	.4646	3.647	108.8	.6720	8.734
#1	18.32	.0760	.0018	3.421	.0013
#2	18.52	.0727	-.0308	3.479	.0012
#3	18.43	.0708	-.0213	3.448	.0012
#4	18.33	.0693	-.0031	3.471	.0012
#5	18.44	.0741	-.0084	3.443	.0010

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	18.18	.0540	.0512	.0075
Stddev	.22	.0013	.0010	.0009
%RSD	1.234	2.405	1.895	12.22
#1	17.80	.0562	.0503	.0067
#2	18.35	.0539	.0506	.0076
#3	18.25	.0529	.0528	.0091
#4	18.34	.0535	.0511	.0071
#5	18.17	.0536	.0513	.0071

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.534
Stddev	.132
%RSD	.67762

#1	19.748
#2	19.485
#3	19.541
#4	19.388
#5	19.510

Method: PRV Sample Name: s98t003385_a Operator: kjt
 Comment: 40501 df
 Run Time: 03/05/99 14:02 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1330	.0706	.0094	17.31	.0011
Stddev	.0021	.0022	.0156	.20	.0002
%RSD	1.614	3.152	165.9	1.141	15.29

#1	.1319	.0722	-.0054	17.19	.0011
#2	.1353	.0669	-.0079	17.49	.0008
#3	.1298	.0709	.0271	17.20	.0011
#4	.1343	.0725	.0122	17.13	.0012
#5	.1335	.0703	.0209	17.56	.0011

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1463	.0541	.0524	.0009
Stddev	.0038	.0021	.0013	.0005
%RSD	2.578	3.844	2.389	52.41

#1	.1406	.0517	.0535	.0004
#2	.1457	.0547	.0511	.0009
#3	.1494	.0527	.0512	.0010
#4	.1460	.0571	.0526	.0005
#5	.1501	.0544	.0537	.0015

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.869
Stddev	.216
%RSD	1.1422

#1	19.151
#2	18.932
#3	18.897
#4	18.557
#5	18.806

HNF-1663 REV. 0

Method: PRV Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/05/99 14:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.32	.0004	-.0071	1.301	-.0002
Stddev	.31	.0005	.0122	.048	.0003
%RSD	1.801	147.8	173.1	3.656	136.7
#1	17.66	.0012	-.0098	1.280	-.0004
#2	17.07	.0002	-.0191	1.277	.0002
#3	16.93	-.0003	-.0159	1.250	-.0002
#4	17.56	.0002	.0113	1.368	-.0001
#5	17.39	.0006	-.0018	1.333	-.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	17.70	.0063	.0103	.0966
Stddev	.29	.0002	.0008	.0021
%RSD	1.661	3.804	7.919	2.203
#1	17.90	.0064	.0100	.0997
#2	17.34	.0062	.0091	.0976
#3	17.63	.0061	.0107	.0942
#4	18.09	.0067	.0108	.0960
#5	17.55	.0063	.0111	.0956

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.664
Stddev	.196
%RSD	1.0504

#1	18.564
#2	18.831
#3	18.865
#4	18.390
#5	18.671

Method: PRV Sample Name: ccv

Operator: kjt

Comment: sn lcs

Run Time: 03/05/99 14:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.4594	-.0005	.0173	20.27	-.0002
Stddev	.4107	.0011	.0126	.17	.0002
%RSD	89.40	217.5	72.44	.8210	156.0
#1	1.143	.0004	.0303	20.11	-.0005
#2	.5280	.0000	.0219	20.51	.0000
#3	.2966	.0004	.0158	20.38	-.0001
#4	.1941	-.0020	-.0031	20.14	-.0003
#5	.1349	-.0012	.0219	20.21	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0092	.0250	.0415	.3886
Stddev	.0007	.0009	.0006	.0079
%RSD	8.050	3.624	1.369	2.042
#1	.0094	.0265	.0409	.3758
#2	.0097	.0251	.0416	.3962
#3	.0097	.0248	.0415	.3868
#4	.0079	.0246	.0413	.3908
#5	.0094	.0242	.0424	.3933

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.288
Stddev	.208
%RSD	1.2758

#1	16.657
#2	16.193
#3	16.240
#4	16.165
#5	16.186

Method: PRV Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/05/99 14:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0010	.0007	-.0103	.0050	.0000
Stddev	.0001	.0012	.0088	.0002	.0002
%RSD	13.99	167.5	85.52	3.862	425.8

#1	.0011	.0010	-.0210	.0049	-.0001
#2	.0009	-.0003	-.0105	.0053	-.0002
#3	.0010	.0017	-.0086	.0051	.0000
#4	.0011	.0020	-.0143	.0049	.0001
#5	.0008	-.0008	.0030	.0049	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0001	-.0001	-.0004
Stddev	.0003	.0004	.0002	.0007
%RSD	55.93	286.5	297.3	193.7

#1	.0009	-.0005	-.0002	.0004
#2	.0000	-.0002	-.0001	-.0003
#3	.0007	-.0003	-.0002	-.0013
#4	.0006	-.0002	.0003	.0002
#5	.0008	.0005	-.0001	-.0010

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.344
Stddev	.060
%RSD	.32941

#1	18.389
#2	18.304
#3	18.261
#4	18.407
#5	18.356

Method: PRV

Sample Name: s98t002391

Operator: kjt

Comment: 11 df

Run Time: 03/05/99 14:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0013	.0022	.0020	.0001
Stddev	.0003	.0013	.0147	.0010	.0002
%RSD	29.12	101.7	678.6	51.91	194.3

#1	.0007	.0001	.0211	.0036	.0001
#2	.0010	.0000	-.0127	.0024	.0000
#3	.0012	.0017	-.0094	.0017	-.0001
#4	.0012	.0032	.0137	.0012	.0003
#5	.0007	.0014	-.0018	.0010	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	.0016	.0028	.0295
Stddev	.0003	.0001	.0005	.0017
%RSD	41.17	8.199	17.83	5.614

#1	.0006	.0018	.0029	.0306
#2	.0006	.0016	.0031	.0312
#3	.0010	.0016	.0031	.0303
#4	.0003	.0016	.0019	.0282
#5	.0005	.0014	.0029	.0274

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.376
Stddev	.661
%RSD	3.5959

#1	17.230
#2	18.912
#3	18.472
#4	18.586
#5	18.682

Method: PRV   Sample Name: s98t002391_d   Operator: kjt
Comment: 11 df
Run Time: 03/05/99 14:45 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0009	.0000	.0020	.0001
Stddev	.0001	.0006	.0171	.0008	.0003
%RSD	10.72	59.54	1224000.	37.83	444.6

#1	.0008	.0018	-.0112	.0030	.0001
#2	.0008	.0011	-.0013	.0024	.0002
#3	.0008	.0007	.0262	.0014	-.0002
#4	.0006	.0005	.0045	.0011	.0003
#5	.0007	.0006	-.0182	.0022	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0005	.0016	.0026	.0260
Stddev	.0002	.0004	.0004	.0013
%RSD	45.38	27.13	14.48	5.126

#1	.0004	.0021	.0028	.0268
#2	.0008	.0011	.0030	.0245
#3	.0004	.0019	.0022	.0279
#4	.0003	.0016	.0022	.0253
#5	.0005	.0012	.0030	.0256

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.414
Stddev	.156
%RSD	.84867

#1	18.491
#2	18.156
#3	18.465
#4	18.392
#5	18.563

Method: PRV Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/05/99 14:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.98	.0015	-.0062	1.253	-.0002
Stddev	.31	.0010	.0114	.056	.0001
%RSD	1.825	67.04	183.3	4.462	45.81

#1	17.45	.0005	-.0158	1.169	-.0002
#2	16.96	.0022	.0072	1.235	-.0001
#3	17.02	.0004	.0038	1.269	-.0001
#4	16.59	.0026	-.0182	1.271	-.0003
#5	16.90	.0020	-.0080	1.320	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	17.57	.0061	.0101	.0973
Stddev	.34	.0004	.0005	.0034
%RSD	1.949	7.114	4.693	3.479

#1	17.90	.0064	.0102	.0990
#2	17.70	.0062	.0104	.1005
#3	17.62	.0054	.0096	.0916
#4	16.99	.0060	.0106	.0976
#5	17.65	.0065	.0096	.0976

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.180
Stddev	.253
%RSD	1.3925

#1	17.861
#2	18.186
#3	18.183
#4	18.566
#5	18.105

HNF-1663 REV. 0

Method: PRV Sample Name: ccv

Operator: kjt

Comment: 19.25 ppb.sn

Run Time: 03/05/99 14:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.5327	.0002	.0135	20.11	.0000
Stddev	.4289	.0005	.0116	.16	.0004
%RSD	80.50	234.0	85.97	.7984	1282.

#1	1.167	.0006	.0134	20.36	-.0001
#2	.7675	.0009	.0056	20.05	.0006
#3	.3686	.0002	.0200	20.11	-.0002
#4	.2168	-.0002	.0288	20.09	.0000
#5	.1437	-.0003	-.0006	19.92	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0103	.0255	.0411	.3798
Stddev	.0014	.0008	.0014	.0037
%RSD	13.98	2.983	3.460	.9846

#1	.0128	.0266	.0419	.3863
#2	.0104	.0253	.0402	.3767
#3	.0096	.0259	.0392	.3795
#4	.0092	.0249	.0428	.3785
#5	.0097	.0248	.0415	.3781

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.313
Stddev	.090
%RSD	.54893

#1	16.313
#2	16.337
#3	16.176
#4	16.315
#5	16.425

HNF-1663 REV. 0

Method: PRV Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/05/99 15:14 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0008	.0016	-.0077	.0066	.0001
Stddev	.0003	.0012	.0168	.0008	.0002
%RSD	31.02	73.60	219.6	12.09	181.5

#1	.0004	.0026	.0142	.0055	-.0002
#2	.0009	-.0001	-.0167	.0062	.0001
#3	.0008	.0026	-.0240	.0072	.0002
#4	.0012	.0009	-.0180	.0075	.0001
#5	.0008	.0020	.0063	.0065	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0007	-.0003	-.0002	-.0023
Stddev	.0002	.0003	.0002	.0008
%RSD	31.74	109.2	65.51	36.16

#1	.0006	-.0006	-.0002	-.0019
#2	.0010	-.0003	-.0005	-.0013
#3	.0004	.0000	-.0003	-.0030
#4	.0006	-.0004	-.0002	-.0033
#5	.0007	.0001	.0000	-.0021

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.861
Stddev	.423
%RSD	2.2413

#1	19.484
#2	18.624
#3	18.816
#4	19.013
#5	18.366

LABCORE Completed Worklist Report for Worklist# 27499

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-6

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000142	1.44e-04	101.408 % Recovery	
1	ICV	0	QMSU-QC U238	QC	0.020	2.08e-02	104.000 % Recovery	
2	ICB	0	QMSU-QC U233	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5		ug/mL
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5		ug/mL
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5		ug/mL
3	SAMPLE	S98T002826	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
3	SAMPLE	S98T002826	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
3	SAMPLE	S98T002826	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
3	SAMPLE	S98T002826	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648	ug/mL
3	SAMPLE	S98T002826	0 D QMSU-D1 U238-D1	LIQUID	N/A	5.285e+00	0.486	ug/mL
4	DUP	S98T002826	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
4	DUP	S98T002826	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
4	DUP	S98T002826	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
4	DUP	S98T002826	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1		RPD
4	DUP	S98T002826	0 D QMSU-D1 U238-D1	LIQUID	5.29e+00	5.29e+00	0.000	RPD
5	SAMPLE	S98T002828	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
5	SAMPLE	S98T002828	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
5	SAMPLE	S98T002828	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
5	SAMPLE	S98T002828	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648	ug/mL
5	SAMPLE	S98T002828	0 D QMSU-D1 U238-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
6	DUP	S98T002828	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
6	DUP	S98T002828	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
6	DUP	S98T002828	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
6	DUP	S98T002828	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1		RPD
6	DUP	S98T002828	0 D QMSU-D1 U238-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
7	SAMPLE	S98T002965	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
7	SAMPLE	S98T002965	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
7	SAMPLE	S98T002965	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486	ug/mL
7	SAMPLE	S98T002965	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648	ug/mL
7	SAMPLE	S98T002965	0 D QMSU-D1 U238-D1	LIQUID	N/A	5.792e-01	0.486	ug/mL
8	DUP	S98T002965	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
8	DUP	S98T002965	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
8	DUP	S98T002965	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1		RPD
8	DUP	S98T002965	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1		RPD
8	DUP	S98T002965	0 D QMSU-D1 U238-D1	LIQUID	5.79e-01	5.31e-01	8.649	RPD
9	SPK-POST	S98T002965	0 D QMSU-D1 U235-D1	LIQUID	0.0001422	1.36e-04	95.640 % Recovery	
9	SPK-POST	S98T002965	0 D QMSU-D1 U238-D1	LIQUID	0.020	1.95e-02	97.500 % Recovery	
10	CCV	0	QMSU-QC U235	QC	0.000142	1.45e-04	102.113 % Recovery	
10	CCV	0	QMSU-QC U238	QC	0.020	2.07e-02	103.500 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27499

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11 CCB	0		QMSU-QC U233	QC	1	<1.20e-5		ug/mL
11 CCB	0		QMSU-QC U234	QC	1	<1.20e-5		ug/mL
11 CCB	0		QMSU-QC U235	QC	1	<1.20e-5		ug/mL
11 CCB	0		QMSU-QC U236	QC	1	<1.60e-5		ug/mL
11 CCB	0		QMSU-QC U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 27499

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul M. Pang 03/01/99
_____
Reviewer Signature Date

12/10/98 06:25

A-0004-1

FILE # 002266.774

Page: 1

LABCORE Data Entry Template for Worklist# 27499

Analyst: B56 Instrument: ICPMS1 Book# 990226-6Method: LA-506-101 Rev/Mod A-0Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	SAMPLE	S98T002826 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: U233-D1, U234-D1, U235-D1, U236-D1, U238-D1				
4	DUP	S98T002826 0 D	@MSU-D1	LIQUID		
5	SAMPLE	S98T002828 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: U233-D1, U234-D1, U235-D1, U236-D1, U238-D1				
6	DUP	S98T002828 0 D	@MSU-D1	LIQUID		
7	SAMPLE	S98T002965 0 D	@MSU-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: U233-D1, U234-D1, U235-D1, U236-D1, U238-D1				
8	DUP	S98T002965 0 D	@MSU-D1	LIQUID		
9	SPK-POST	S98T002965 0 D	@MSU-D1	LIQUID		
10	CCV		@MSU-QC	QC		
11	CCB		@MSU-QC	QC		

Final page for worklist # 27499

Analyst Signature

Date

Analyst Signature

Date

S98T002826

26-D

S98T002828

28-D

S98T002865

65-D

65-A

Data Entry Comments:

POST spike - .02ml of 10ppm U500 (20ppb)

SAMS - 982826, 982828, 982865

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2431

Analysis Report

02/26/99 20:48:26

page 1

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 02/26/99

20:47 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	197200.
Stddev	1346.
%RSD	.6828

#1	197900.
#2	196500.
#3	197000.
#4	195500.
#5	198500.
#6	195000.
#7	197600.
#8	196900.
#9	199500.
#10	197300.

*bjg**By-OS**FSOU**598700 2826**2828**2865**FILE# 002266.TFT**W.L.# 27488*

uploaded + validated by
Saul R. Pang
03/01/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2432 TO 2439.

2432

Analysis Report

02/26/99 20:50:06

page 1

Method: ISOU Sample Name: blank

Operator: bjg.

Comment: background

Run Time: 02/26/99 20:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0013	-.0015	-.0022	-.0022	-.0015
Stddev	.0010	.0011	.0005	.0004	.0012
%RSD	77.56	75.19	23.21	16.65	83.82

#1	-.0018	-.0032	-.0029	-.0018	-.0027
#2	-.0019	-.0006	-.0020	-.0026	-.0016
#3	.0005	-.0018	-.0021	-.0021	-.0023
#4	-.0020	-.0005	-.0024	-.0020	-.0012
#5	-.0015	-.0013	-.0015	-.0026	.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	13.577
Stddev	.103
%RSD	.75998

#1	13.596
#2	13.614
#3	13.436
#4	13.527
#5	13.713

2433

Analysis Report

02/26/99 20:53:34

page 1

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 20:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0016	.0000	.1438	.0006	20.77
Stddev	.0008	.001	.0030	.0011	.22
%RSD	52.26	6943.	2.108	198.8	1.082
#1	-.0025	-.0021	.1482	.0012	20.89
#2	-.0011	-.0009	.1409	.0019	20.67
#3	-.0005	.0011	.1442	.0010	20.43
#4	-.0023	.0010	.1447	-.0002	20.82
#5	-.0017	.0008	.1410	-.0010	21.02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.415
Stddev	.091
%RSD	.67504

#1	13.254
#2	13.441
#3	13.460
#4	13.452
#5	13.467

2434

Analysis Report

02/26/99 21:00:58

page 1

Method: ISOU

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 20:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0022	-.0005	-.0021	-.0022	-.0021
Stddev	.0005	.0009	.0013	.0004	.0013
%RSD	23.81	190.9	60.25	19.20	60.48
#1	-.0022	-.0007	-.0007	-.0026	-.0030
#2	-.0030	-.0019	-.0034	-.0021	-.0001
#3	-.0024	-.0003	-.0034	-.0015	-.0017
#4	-.0017	.0003	-.0022	-.0025	-.0028
#5	-.0018	.0003	-.0010	-.0024	-.0031

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.380
Stddev	.031
%RSD	.23368

#1	13.345
#2	13.386
#3	13.401
#4	13.416
#5	13.350

2435

Analysis Report

02/26/99 21:04:49

page 1

Method: ISOU Sample Name: s98t002826 Operator: bjg
 Comment: 40501 df
 Run Time: 02/26/99 21:03 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0022	-.0007	-.0012	-.0028	.1305
Stddev	.0005	.0014	.0011	.0010	.0035
%RSD	23.09	195.3	95.70	37.06	2.701
#1	-.0027	-.0021	-.0010	-.0014	.1290
#2	-.0014	-.0011	-.0017	-.0040	.1284
#3	-.0026	-.0009	.0002	-.0021	.1321
#4	-.0022	.0017	-.0007	-.0033	.1272
#5	-.0025	-.0013	-.0028	-.0031	.1360

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.465
Stddev	.138
%RSD	1.0264

#1	13.352
#2	13.705
#3	13.434
#4	13.434
#5	13.401

2436

Analysis Report

02/26/99 21:07:36

page 1

Method: ISOU

Sample Name: s98t002826_d

Operator: bjg

Comment: 40501 df

Run Time: 02/26/99

21:06 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0028	-.0015	-.0011	-.0031	.1305
Stddev	.0007	.0007	.0004	.0012	.0028
%RSD	25.92	45.89	38.19	37.63	2.144
#1	-.0020	-.0012	-.0007	-.0037	.1313
#2	-.0032	-.0018	-.0015	-.0012	.1339
#3	-.0020	-.0022	-.0008	-.0028	.1319
#4	-.0032	-.0017	-.0011	-.0041	.1278
#5	-.0035	-.0004	-.0017	-.0038	.1274

Int. Std.

	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.914
Stddev	.274
%RSD	1.9689

#1	13.533
#2	13.835
#3	13.839
#4	14.173
#5	14.189

2437

Analysis Report

02/26/99 21:09:44

page 1

Method: ISOU

Sample Name: s98t002828

Operator: bjg

Comment: 40501 df

Run Time: 02/26/99 21:08 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0025	-.0011	-.0021	-.0026	.0072
Stddev	.0012	.0006	.0007	.0005	.0017
%RSD	47.92	51.96	32.77	19.05	24.23
#1	-.0005	-.0015	-.0018	-.0028	.0077
#2	-.0035	-.0012	-.0016	-.0032	.0088
#3	-.0024	-.0006	-.0033	-.0029	.0042
#4	-.0026	-.0004	-.0020	-.0019	.0079
#5	-.0032	-.0017	-.0018	-.0025	.0073

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.084
Stddev	.063
%RSD	.44628

#1	14.115
#2	14.065
#3	14.112
#4	13.983
#5	14.143

2438

Analysis Report

02/26/99 21:13:37

page 1

Method: ISOU   Sample Name: s98t002828_d   Operator: bjg
 Comment: 40501 df
 Run Time: 02/26/99 21:12 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0026	-.0016	-.0028	-.0023	.0068
Stddev	.0013	.0009	.0008	.0008	.0017
%RSD	50.88	54.66	28.73	36.41	24.81
#1	-.0026	-.0016	-.0019	-.0027	.0054
#2	-.0044	-.0027	-.0023	-.0023	.0093
#3	-.0016	-.0003	-.0031	-.0017	.0069
#4	-.0011	-.0020	-.0030	-.0013	.0073
#5	-.0034	-.0016	-.0040	-.0034	.0052

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.143
Stddev	.178
%RSD	1.2568

#1	14.256
#2	14.001
#3	14.227
#4	13.909
#5	14.319

Analysis Report

02/26/99 21:18:26

page 1

Method: ISOU

Sample Name: s98t002965

Operator: bjg

Comment: 40501 df

Run Time: 02/26/99 21:17 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0020	-.0022	-.0029	-.0032	.0143
Stddev	.0007	.0007	.0006	.0006	.0014
%RSD	34.96	31.42	21.57	17.31	9.698
#1	-.0022	-.0016	-.0025	-.0027	.0158
#2	-.0009	-.0033	-.0032	-.0041	.0127
#3	-.0021	-.0022	-.0024	-.0035	.0136
#4	-.0020	-.0020	-.0028	-.0029	.0138
#5	-.0027	-.0019	-.0039	-.0031	.0158

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	14.065
Stddev	.107
%RSD	.76359

#1	13.982
#2	14.077
#3	14.115
#4	13.942
#5	14.211

2440

Method: ISOU   Sample Name: s98t002965_d   Operator: bjg
Comment: 40501 df
Run Time: 02/26/99 21:19 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0020	-.0026	-.0026	-.0022	.0131
Stddev	.0015	.0010	.0005	.0010	.0012
%RSD	74.43	38.85	20.43	46.95	9.326
#1	-.0008	-.0038	-.0030	-.0027	.0132
#2	-.0010	-.0013	-.0025	-.0005	.0129
#3	-.0035	-.0032	-.0031	-.0025	.0112
#4	-.0009	-.0025	-.0023	-.0021	.0135
#5	-.0039	-.0020	-.0019	-.0030	.0146

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	14.193
Stddev	.233
%RSD	1.6451

#1	14.419
#2	13.854
#3	14.293
#4	14.055
#5	14.345

2441

Method: ISOU   Sample Name: s98t002965_a   Operator: bjg
Comment: 40501 df
Run Time: 02/26/99 21:22 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0024	-.0006	.1362	-.0006	19.52
Stddev	.0009	.0017	.0031	.0012	.33
%RSD	37.06	275.0	2.254	204.4	1.692
#1	-.0014	-.0017	.1369	.0011	19.19
#2	-.0036	.0007	.1380	-.0009	19.14
#3	-.0024	.0007	.1342	-.0004	19.84
#4	-.0027	.0003	.1398	-.0021	19.65
#5	-.0016	-.0031	.1321	-.0005	19.78

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.325
Stddev	.130
%RSD	.90433

#1	14.391
#2	14.345
#3	14.097
#4	14.381
#5	14.410

Analysis Report

02/26/99 21:25:54

page 1

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 21:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_ [233]	U_ [234]	U_ [235]	U_ [236]	U_ [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0023	-.0002	.1445	-.0008	20.73
Stddev	.0007	.0010	.0040	.0011	.31
%RSD	30.38	649.0	2.756	140.0	1.504
#1	-.0014	.0005	.1437	-.0019	21.06
#2	-.0022	.0005	.1453	-.0004	20.52
#3	-.0029	-.0010	.1382	-.0005	20.51
#4	-.0031	-.0015	.1482	-.0019	20.48
#5	-.0019	.0007	.1474	.0007	21.09

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.265
Stddev	.177
%RSD	1.2427

#1	13.990
#2	14.293
#3	14.456
#4	14.367
#5	14.217

2443

Analysis Report

02/26/99 21:32:42

page 1

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 21:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0022	-.0026	-.0021	-.0033	-.0013
Stdddev	.0006	.0006	.0009	.0005	.0014
%RSD	29.29	24.33	41.57	14.55	104.0
#1	-.0021	-.0025	-.0024	-.0039	-.0003
#2	-.0026	-.0028	-.0015	-.0032	-.0000
#3	-.0012	-.0018	-.0009	-.0036	-.0007
#4	-.0020	-.0035	-.0030	-.0029	-.0026
#5	-.0029	-.0023	-.0026	-.0027	-.0030

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.556
Stdddev	.146
%RSD	1.0021

#1	14.604
#2	14.317
#3	14.544
#4	14.707
#5	14.609

2444

LABCORE Completed Worklist Report for Worklist# 27500

Analyst: kjt

Instrument: ICPMS1

Book#: 48B57

Method: LA-506-101 Rev/Mod A-0

49B57

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSPR-QC TC99	QC	0.0205	1.90e-02	92.683 % Recovery	
1	ICV	0	QMSPR-QC SN117	QC	0.01925	2.15e-02	111.688 % Recovery	
1	ICV	0	QMSPR-QC CS133	QC	0.02000	1.97e-02	98.500 % Recovery	
2	ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3	SAMPLE	S98T002829 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	5.743e+00	4.06e-002	ug/mL
3	SAMPLE	S98T002829 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
3	SAMPLE	S98T002829 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
3	SAMPLE	S98T002829 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	6.709e+00	4.06e-002	ug/mL
3	SAMPLE	S98T002829 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.458e+00	4.06e-002	ug/mL
3	SAMPLE	S98T002829 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.420e+00	4.06e-002	ug/mL
4	DUP	S98T002829 0 D	QMSPR-D1 TC99-D1	LIQUID	5.74e+00	5.70e+00	0.699	RPD
4	DUP	S98T002829 0 D	QMSPR-D1 SN117-D1	LIQUID	<3.04e0	<3.05e0		RPD
4	DUP	S98T002829 0 D	QMSPR-D1 SN126-D1	LIQUID	<3.04e0	<3.05e0		RPD
4	DUP	S98T002829 0 D	QMSPR-D1 CS133-D1	LIQUID	6.71e+00	6.47e+00	3.642	RPD
4	DUP	S98T002829 0 D	QMSPR-D1 CS135-D1	LIQUID	2.46e+00	2.44e+00	0.816	RPD
4	DUP	S98T002829 0 D	QMSPR-D1 CS137-D1	LIQUID	2.42e+00	2.44e+00	0.823	RPD
5	SAMPLE	S98T002966 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	4.945e+00	4.06e-002	ug/mL
5	SAMPLE	S98T002966 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
5	SAMPLE	S98T002966 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
5	SAMPLE	S98T002966 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	5.532e+00	4.06e-002	ug/mL
5	SAMPLE	S98T002966 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.065e+00	4.06e-002	ug/mL
5	SAMPLE	S98T002966 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.012e+00	4.06e-002	ug/mL
6	DUP	S98T002966 0 D	QMSPR-D1 TC99-D1	LIQUID	4.95e+00	4.83e+00	2.454	RPD
6	DUP	S98T002966 0 D	QMSPR-D1 SN117-D1	LIQUID	<3.04e0	<3.05e0		RPD
6	DUP	S98T002966 0 D	QMSPR-D1 SN126-D1	LIQUID	<3.04e0	<3.05e0		RPD
6	DUP	S98T002966 0 D	QMSPR-D1 CS133-D1	LIQUID	5.53e+00	5.57e+00	0.721	RPD
6	DUP	S98T002966 0 D	QMSPR-D1 CS135-D1	LIQUID	2.06e+00	2.03e+00	1.467	RPD
6	DUP	S98T002966 0 D	QMSPR-D1 CS137-D1	LIQUID	2.01e+00	1.97e+00	2.010	RPD
7	SAMPLE	S98T002967 0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	6.117e+00	4.06e-002	ug/mL
7	SAMPLE	S98T002967 0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
7	SAMPLE	S98T002967 0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	3.045e+00	3.045	ug/mL
7	SAMPLE	S98T002967 0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	7.020e+00	4.06e-002	ug/mL
7	SAMPLE	S98T002967 0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.592e+00	4.06e-002	ug/mL
7	SAMPLE	S98T002967 0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	2.491e+00	4.06e-002	ug/mL
8	DUP	S98T002967 0 D	QMSPR-D1 TC99-D1	LIQUID	6.12e+00	6.33e+00	3.373	RPD
8	DUP	S98T002967 0 D	QMSPR-D1 SN117-D1	LIQUID	<3.04e0	<3.05e0		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27500

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
8 DUP	S98T002967	0 D	QMSPR-D1 SN126-D1	LIQUID	<3.04e0	<3.05e0		RPD
8 DUP	S98T002967	0 D	QMSPR-D1 CS133-D1	LIQUID	7.02e+00	6.95e+00	1.002	RPD
8 DUP	S98T002967	0 D	QMSPR-D1 CS135-D1	LIQUID	2.59e+00	2.57e+00	0.775	RPD
8 DUP	S98T002967	0 D	QMSPR-D1 CS137-D1	LIQUID	2.49e+00	2.61e+00	4.706	RPD
9 SPK-POST	S98T002967	0 D	QMSPR-D1 TC99-D1	LIQUID	0.0219	2.14e-02	97.717	% Recovery
9 SPK-POST	S98T002967	0 D	QMSPR-D1 SN117-D1	LIQUID	0.0193	1.88e-02	97.409	% Recovery
9 SPK-POST	S98T002967	0 D	QMSPR-D1 CS133-D1	LIQUID	0.0201	2.01e-02	100.000	% Recovery
10 CCV		0	QMSPR-QC TC99	QC	0.0205	1.91e-02	93.171	% Recovery
10 CCV		0	QMSPR-QC SN117	QC	0.01925	2.13e-02	110.649	% Recovery
10 CCV		0	QMSPR-QC CS133	QC	0.02000	1.90e-02	95.000	% Recovery
11 CCB		0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
11 CCB		0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
11 CCB		0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
11 CCB		0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
11 CCB		0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
11 CCB		0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27500

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul H. Parag 03/08/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27500

Analyst: KST Instrument: ICPMS1 Book# 48857
Method: LA-506-101 Rev/Mod A-0 49857
Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSPR-QC	QC		
2	ICB		@MSPR-QC	QC		
3	SAMPLE	S98T002829 0 D	@MSPR-D1	LIQUID	98000437	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1				
4	DUP	S98T002829 0 D	@MSPR-D1	LIQUID		
5	SAMPLE	S98T002966 0 D	@MSPR-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1				
6	DUP	S98T002966 0 D	@MSPR-D1	LIQUID		
7	SAMPLE	S98T002967 0 D	@MSPR-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1				
8	DUP	S98T002967 0 D	@MSPR-D1	LIQUID		
9	SPK-POST	S98T002967 0 D	@MSPR-D1	LIQUID		
10	CCV		@MSPR-QC	QC		
11	CCB		@MSPR-QC	QC		

Final page for worklist # 27500

K. S. Romo 3-7-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

Uploaded + Validated by:
Saul M. Pang
03/08/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/07/99 13:41 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem In[115]
 Line 115/pulse
 Units Cts/S
 Avg 176300.
 Stddev 854.
 %RSD .4847

#1 176700.
 #2 175500.
 #3 176000.
 #4 177500.
 #5 175200.
 #6 176400.
 #7 174900.
 #8 176400.
 #9 176600.
 #10 177300.

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2448 TO 2449

file pr0307a.txt

S98T00 2829

2829-d

2966

2966-d

2967

2967-d

2967-a

2967-a

1 .05
 25-10-1-10 20301 df
 KST 307-49

Scans

pr 2829

pr 2966

pr 2967

KJ Thondike 3299

Spike .1ml of msprv spike 4487
 .25ml of Open Sn (019-1)
 2448

Method: PRV Sample Name: icv

Operator: kjt

Comment: ms prv lcs

Run Time: 03/07/99 13:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.95	.0009	-.0023	1.578	.0002
Stddev	.14	.0005	.0097	.074	.0004
%RSD	.7125	53.72	422.1	4.708	168.7
#1	19.15	.0009	-.0082	1.483	.0005
#2	18.78	.0007	-.0059	1.524	.0006
#3	18.97	.0003	.0133	1.587	-.0004
#4	18.88	.0010	.0003	1.639	.0001
#5	18.97	.0016	-.0110	1.657	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.70	.0074	.0125	.1136
Stddev	.20	.0002	.0009	.0034
%RSD	1.037	3.148	7.331	3.014
#1	20.04	.0076	.0131	.1130
#2	19.68	.0073	.0124	.1139
#3	19.61	.0071	.0114	.1106
#4	19.69	.0076	.0138	.1193
#5	19.50	.0073	.0120	.1114

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.689
Stddev	.118
%RSD	.66540

#1	17.504
#2	17.832
#3	17.697
#4	17.695
#5	17.719

Method: PRV Sample Name: icv

Operator: kjt

Comment: 19.25 ppb Sn

Run Time: 03/07/99 13:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0432	-.0001	.0200	21.46	.0003
Stddev	.0178	.0007	.0175	.35	.0003
%RSD	41.28	1047.	87.58	1.628	112.1
#1	.0689	.0009	.0052	21.27	-.0001
#2	.0530	-.0005	-.0022	21.78	.0004
#3	.0380	-.0008	.0353	21.87	.0008
#4	.0323	.0003	.0258	21.10	.0004
#5	.0240	-.0002	.0357	21.25	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0120	.0286	.0470	.4326
Stddev	.0008	.0021	.0015	.0155
%RSD	6.971	7.480	3.283	3.572
#1	.0117	.0272	.0462	.4198
#2	.0123	.0288	.0469	.4506
#3	.0133	.0320	.0495	.4483
#4	.0112	.0265	.0469	.4239
#5	.0114	.0286	.0454	.4205

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.598
Stddev	.378
%RSD	2.4238

#1	16.007
#2	15.443
#3	15.026
#4	15.721
#5	15.792

Method: PRV

Sample Name: ich

Operator: kjt

Comment: 2% nitric

Run Time: 03/07/99 14:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0007	-.0126	.0052	.0002
Stddev	.0003	.0005	.0126	.0012	.0002
%RSD	94.70	73.32	100.0	22.48	85.95
#1	-.0005	.0007	-.0290	.0053	.0004
#2	-.0003	.0012	-.0037	.0032	.0001
#3	-.0005	.0006	.0012	.0054	.0003
#4	.0002	.0013	-.0217	.0061	.0003
#5	-.0007	.0000	-.0094	.0060	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0010	.0000	.0000	-.0011
Stddev	.0005	.0002	.0001	.0009
%RSD	45.99	922.7	621.0	80.40
#1	.0017	.0001	.0002	-.0014
#2	.0005	.0000	-.0001	-.0021
#3	.0008	-.0003	.0000	-.0016
#4	.0013	.0000	.0000	.0002
#5	.0008	.0003	.0000	-.0006

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	17.125
Stddev	.246
%RSD	1.4374

#1	17.538
#2	16.956
#3	17.137
#4	17.069
#5	16.925

Method: PRV

Sample Name: s98t002829

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.2829	.1525	-.0069	.0940	.0032
Stddev	.0064	.0036	.0132	.0077	.0004
%RSD	2.248	2.359	191.3	8.195	13.36
#1	.2863	.1560	-.0268	.0948	.0039
#2	.2841	.1547	-.0050	.0891	.0033
#3	.2725	.1502	-.0073	.0896	.0028
#4	.2824	.1540	.0101	.0895	.0030
#5	.2892	.1473	-.0055	.1072	.0030

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.3305	.1211	.1192	.0052
Stddev	.0097	.0019	.0060	.0011
%RSD	2.943	1.598	5.065	21.39
#1	.3305	.1240	.1211	.0057
#2	.3463	.1205	.1192	.0044
#3	.3256	.1201	.1127	.0048
#4	.3203	.1190	.1149	.0041
#5	.3300	.1221	.1282	.0068

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.750
Stddev	.546
%RSD	3.2595

#1	16.306
#2	16.804
#3	17.298
#4	17.252
#5	16.088

Method: PRV Sample Name: s98t002829_d Operator: kjt
 Comment: 20301 df
 Run Time: 03/07/99 14:34 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.2809	.1438	.0068	.0958	.0030
Stddev	.0083	.0072	.0080	.0033	.0005
%RSD	2.965	5.022	118.6	3.481	15.62

#1	.2899	.1440	.0091	.0961	.0026
#2	.2767	.1468	-.0026	.0977	.0036
#3	.2691	.1358	.0001	.0974	.0025
#4	.2814	.1384	.0172	.0900	.0033
#5	.2872	.1541	.0100	.0978	.0032

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.3187	.1200	.1201	-.0033
Stddev	.0101	.0061	.0072	.0015
%RSD	3.177	5.112	5.978	45.96

#1	.3040	.1202	.1205	-.0031
#2	.3192	.1143	.1104	-.0054
#3	.3153	.1140	.1159	-.0039
#4	.3312	.1231	.1286	-.0028
#5	.3238	.1284	.1249	-.0013

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.252
Stddev	.216
%RSD	1.3278

#1	16.050
#2	16.326
#3	16.577
#4	16.067
#5	16.240

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002966

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.2436	.1268	-.0009	.0856	.0026
Stddev	.0063	.0033	.0153	.0023	.0003
%RSD	2.574	2.631	1774.	2.628	11.07

#1	.2452	.1273	.0001	.0862	.0028
#2	.2518	.1320	-.0048	.0882	.0024
#3	.2401	.1240	.0081	.0843	.0022
#4	.2352	.1271	-.0241	.0870	.0030
#5	.2457	.1238	.0164	.0825	.0027

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2725	.1017	.0991	.0033
Stddev	.0053	.0029	.0038	.0013
%RSD	1.953	2.835	3.804	38.40

#1	.2685	.1044	.1040	.0033
#2	.2654	.1005	.0992	.0053
#3	.2762	.0973	.0939	.0019
#4	.2778	.1023	.0974	.0031
#5	.2747	.1039	.1009	.0029

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.607
Stddev	.258
%RSD	1.6507

#1	15.804
#2	15.544
#3	15.870
#4	15.599
#5	15.216

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002966_d

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.2378	.1238	-.0074	.0806	.0026
Stddev	.0125	.0068	.0115	.0057	.0003
%RSD	5.254	5.464	155.9	7.032	11.57

#1	.2396	.1265	-.0074	.0752	.0023
#2	.2376	.1157	-.0120	.0783	.0027
#3	.2192	.1338	-.0234	.0762	.0023
#4	.2381	.1217	.0070	.0874	.0030
#5	.2544	.1213	-.0010	.0859	.0026

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2745	.1001	.0970	.0034
Stddev	.0107	.0036	.0030	.0014
%RSD	3.888	3.633	3.135	42.45

#1	.2758	.0964	.0963	.0035
#2	.2665	.1019	.0924	.0035
#3	.2614	.0974	.0976	.0010
#4	.2815	.0992	.1007	.0048
#5	.2876	.1054	.0980	.0040

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.435
Stddev	.729
%RSD	4.7223

#1	15.632
#2	15.264
#3	16.482
#4	15.336
#5	14.460

Method: PRV

Sample Name: s98t002967

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.3013	.1545	-.0033	.1038	.0032
Stddev	.0153	.0172	.0161	.0016	.0006
%RSD	5.077	11.11	493.1	1.495	19.61
#1	.3281	.1763	-.0221	.1023	.0030
#2	.2923	.1686	.0052	.1048	.0027
#3	.2911	.1490	-.0180	.1029	.0034
#4	.2991	.1372	.0028	.1060	.0042
#5	.2960	.1414	.0157	.1030	.0027

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.3458	.1277	.1227	-.0050
Stddev	.0167	.0031	.0039	.0009
%RSD	4.839	2.421	3.182	18.33
#1	.3391	.1263	.1170	-.0064
#2	.3202	.1314	.1255	-.0038
#3	.3581	.1305	.1240	-.0054
#4	.3617	.1242	.1264	-.0048
#5	.3497	.1262	.1206	-.0048

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.986
Stddev	.557
%RSD	3.7161

#1	15.617
#2	15.290
#3	15.155
#4	14.198
#5	14.670

Analysis Report

HNF-1663 REV. 0

03/07/99 14:47:38

page 1

Method: PRV

Sample Name: s98t002967_d

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.3119	.1518	-.0070	.1016	.0034
Stddev	.0113	.0041	.0159	.0050	.0003
%RSD	3.634	2.693	228.3	4.912	8.337

#1	.3311	.1492	-.0186	.1045	.0039
#2	.3052	.1579	.0014	.1054	.0033
#3	.3131	.1501	.0132	.1048	.0034
#4	.3035	.1540	-.0268	.0937	.0034
#5	.3066	.1479	-.0040	.0996	.0032

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.3425	.1266	.1284	-.0038
Stddev	.0181	.0014	.0054	.0019
%RSD	5.275	1.070	4.168	50.85

#1	.3655	.1277	.1325	-.0005
#2	.3488	.1278	.1357	-.0039
#3	.3199	.1246	.1242	-.0047
#4	.3290	.1272	.1241	-.0055
#5	.3492	.1259	.1255	-.0042

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.690
Stddev	.542
%RSD	3.6912

#1	13.816
#2	14.814
#3	15.050
#4	15.194
#5	14.577

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002967_a

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	21.71	.1419	.0066	3.663	.0034
Stddev	.59	.0029	.0155	.101	.0007
%RSD	2.725	2.043	234.8	2.748	19.52
#1	22.24	.1457	.0154	3.693	.0029
#2	21.38	.1414	.0059	3.499	.0030
#3	21.03	.1437	.0144	3.648	.0030
#4	21.48	.1406	.0174	3.767	.0045
#5	22.41	.1381	.0200	3.705	.0036

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	20.42	.1287	.1236	.0088
Stddev	1.00	.0038	.0026	.0014
%RSD	4.910	2.971	2.115	15.98
#1	20.11	.1255	.1216	.0068
#2	19.16	.1318	.1222	.0103
#3	20.00	.1258	.1215	.0083
#4	21.09	.1337	.1261	.0099
#5	21.72	.1265	.1268	.0086

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.690
Stddev	.242
%RSD	1.7672

#1	13.735
#2	14.087
#3	13.543
#4	13.475
#5	13.608

HNF-1663 REV.0

Method: PRV

Sample Name: s98t002967_a

Operator: kjt

Comment: 20301 df

Run Time: 03/07/99 14:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.3060	.1459	.0113	18.83	.0031
Stddev	.0111	.0035	.0185	.41	.0005
%RSD	3.629	2.421	163.7	2.162	16.33
#1	.2967	.1444	.0400	18.92	.0034
#2	.3241	.1520	-.0024	18.39	.0032
#3	.3062	.1449	.0109	18.52	.0031
#4	.2972	.1451	-.0072	18.87	.0022
#5	.3057	.1430	.0152	19.44	.0034

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.3495	.1292	.1263	-.0033
Stddev	.0145	.0046	.0062	.0009
%RSD	4.143	3.584	4.913	28.81
#1	.3405	.1288	.1261	-.0038
#2	.3639	.1311	.1223	-.0021
#3	.3472	.1213	.1227	-.0039
#4	.3316	.1329	.1234	-.0042
#5	.3645	.1318	.1371	-.0024

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.788
Stddev	.398
%RSD	2.6933

#1	14.987
#2	14.304
#3	15.066
#4	15.167
#5	14.414

page 1

Analysis Report

HNF-1663 REV. 0

03/07/99 15:02:18

page 1

Method: PRV

Sample Name: ~~icv~~

Operator: kjt

Comment: msprv lcs

Run Time: 03/07/99 15:00 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc [99]	Ru [102]	Te [125]	Sn [117]	Sn [126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.05	-.0001	-.0057	1.295	.0002
Stddev	.81	.0010	.0124	.108	.0003
%RSD	4.268	802.8	217.1	8.304	132.3
#1	19.02	-.0016	-.0188	1.154	.0005
#2	19.57	.0005	-.0125	1.232	-.0002
#3	18.20	.0011	.0139	1.314	.0004
#4	20.11	-.0002	-.0026	1.340	.0004
#5	18.33	-.0004	-.0086	1.437	.0001

Elem	Cs [133]	Cs [135]	Cs [137]	Ba [138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.00	.0075	.0126	.1148
Stddev	.26	.0008	.0004	.0060
%RSD	1.371	11.23	3.440	5.256
#1	19.43	.0085	.0125	.1196
#2	18.74	.0077	.0122	.1087
#3	18.88	.0070	.0133	.1219
#4	18.99	.0077	.0126	.1151
#5	18.97	.0063	.0123	.1089

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.027
Stddev	.422
%RSD	2.8080

#1	14.353
#2	14.925
#3	15.276
#4	15.135
#5	15.443

Analysis Report

HNF-1663 REV. 0

03/07/99 15:15:38

page 1

Method: PRV

Sample Name: *low CCV*

Operator: kjt

Comment: 19.25 ppb Sn

Run Time: 03/07/99 15:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0132	-.0044	.0075	21.27	.0006
Stddev	.0086	.0006	.0226	1.70	.0002
%RSD	65.16	12.66	300.0	7.970	34.49

#1	.0279	-.0044	.0290	20.15	.0009
#2	.0132	-.0045	.0288	24.08	.0006
#3	.0100	-.0039	-.0032	21.63	.0005
#4	.0073	-.0040	-.0243	20.44	.0004
#5	.0074	-.0053	.0075	20.03	.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0124	.0267	.0441	.4122
Stddev	.0005	.0012	.0040	.0406
%RSD	4.081	4.644	9.032	9.859

#1	.0123	.0269	.0509	.4389
#2	.0117	.0275	.0442	.4694
#3	.0126	.0245	.0406	.3751
#4	.0125	.0273	.0427	.3975
#5	.0131	.0273	.0423	.3800

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.0396
Stddev	.2568
%RSD	3.1946

#1	8.0731
#2	7.7975
#3	8.4258
#4	8.0908
#5	7.8108

HNF-1663 REV. 0

Method: PRV Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/07/99 15:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0005	.0011	-.0039	.0052	.0002
Stddev	.0003	.0010	.0147	.0008	.0002
%RSD	60.45	93.45	375.5	15.96	79.24

#1	-.0001	.0001	-.0176	.0061	.0001
#2	-.0003	.0027	-.0043	.0048	.0002
#3	-.0009	.0014	-.0140	.0041	.0000
#4	-.0006	.0007	.0201	.0050	.0005
#5	-.0004	.0005	-.0039	.0059	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0001	-.0001	.0000	.0001
Stddev	.0003	.0003	.000	.0005
%RSD	220.3	287.8	3055.	601.4

#1	.0006	.0004	.0003	.0007
#2	-.0001	-.0004	.0004	.0000
#3	-.0001	-.0003	-.0003	-.0005
#4	.0002	.0001	.0000	.0004
#5	.0000	-.0003	-.0004	-.0002

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.414
Stddev	.138
%RSD	.79128

#1	17.492
#2	17.475
#3	17.532
#4	17.188
#5	17.382

LABCORE Completed Worklist Report for Worklist# 27501

Analyst: bjg

Instrument: ICPMS1

Book#: 990301-1

Method: LA-506-101 Rev/Mod A0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000142	1.53e-04	107.746 % Recovery	
1	ICV	0	QMSU-QC U238	QC	0.020	2.24e-02	112.000 % Recovery	
2	ICB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	SAMPLE	S98T002829	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002829	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002829	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T002829	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
3	SAMPLE	S98T002829	0 D QMSU-D1 U238-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
4	DUP	S98T002829	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002829	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002829	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T002829	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
4	DUP	S98T002829	0 D QMSU-D1 U238-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
5	SAMPLE	S98T002966	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002966	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002966	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
5	SAMPLE	S98T002966	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
5	SAMPLE	S98T002966	0 D QMSU-D1 U238-D1	LIQUID	N/A	< 5.468e-01	0.486 ug/mL	
6	DUP	S98T002966	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002966	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002966	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
6	DUP	S98T002966	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
6	DUP	S98T002966	0 D QMSU-D1 U238-D1	LIQUID	5.47e-01	5.43e-01	0.734 RPD	
7	SAMPLE	S98T002967	0 D QMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002967	0 D QMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002967	0 D QMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
7	SAMPLE	S98T002967	0 D QMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
7	SAMPLE	S98T002967	0 D QMSU-D1 U238-D1	LIQUID	N/A	< 5.225e-01	0.486 ug/mL	
8	DUP	S98T002967	0 D QMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002967	0 D QMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002967	0 D QMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
8	DUP	S98T002967	0 D QMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
8	DUP	S98T002967	0 D QMSU-D1 U238-D1	LIQUID	5.22e-01	5.71e-01	8.966 RPD	
9	SPK-POST	S98T002967	0 D QMSU-D1 U235-D1	LIQUID	0.000142	1.45e-04	102.113 % Recovery	
9	SPK-POST	S98T002967	0 D QMSU-D1 U238-D1	LIQUID	0.020	2.11e-02	105.500 % Recovery	
10	CCV	0	QMSU-QC U235	QC	0.000142	1.57e-04	110.563 % Recovery	
10	CCV	0	QMSU-QC U238	QC	0.020	2.27e-02	113.500 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27501

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	CCB	0			QMSU-QC U233	QC	1	<1.20e-5		ug/mL
11	CCB	0			QMSU-QC U234	QC	1	<1.20e-5		ug/mL
11	CCB	0			QMSU-QC U235	QC	1	<1.20e-5		ug/mL
11	CCB	0			QMSU-QC U236	QC	1	<1.60e-5		ug/mL
11	CCB	0			QMSU-QC U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 27501

_____
Analyst Signature   Date

_____
Analyst Signature   Date

Sally M. Pang 03/02/99
Reviewer Signature   Date

03/01/99 15:56

ws2

FILE # 00301A-721

Page: 1

LABCORE Data Entry Template for Worklist# 27501

Analyst: BJG Instrument: ICPMS1 Book# 990301-1Method: LA-506-101 Rev/Mod A-0Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	SAMPLE	S98T002829 0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
4	DUP	S98T002829 0 D	@MSU-D1	LIQUID		
5	SAMPLE	S98T002966 0 D	@MSU-D1	LIQUID	98000509	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
6	DUP	S98T002966 0 D	@MSU-D1	LIQUID		
7	SAMPLE	S98T002967 0 D	@MSU-D1	LIQUID	98000509	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
8	DUP	S98T002967 0 D	@MSU-D1	LIQUID		
9	SPK-POST	S98T002967 0 D	@MSU-D1	LIQUID		
10	CCV		@MSU-QC	QC		
11	CCB		@MSU-QC	QC		

Final page for worklist # 27501

Signature [Signature] Date 3-1-99

Signature [Signature] Date 3-1-99

S98T00 2829
2829-D
S98T00 2966
2966-D
S98T00 2967
2967-D
Data Entry Comments: 2967-A

1-10-1 - 04-10-1 D.C.
05-10-1 M 40501
3-1-99

POST SPIKE = .02 ml of 10ppm U STD in 10 ml 2% HNO₃. (20/4/96)

SCANS = 982829, 982966, 982967

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2465

Analysis Report

03/01/99 17:08:05

page 1

Method: MS_TUNE Sample Name: ms-tune Operator: bjg
Comment:
Run Time: 03/01/99 17:07 Type: Unk Mode: INT Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	219800.
Stddev	1643.
%RSD	.7474

#1	221800.
#2	218800.
#3	219200.
#4	221200.
#5	220300.
#6	218500.
#7	217700.
#8	221500.
#9	221300.
#10	217400.

By-105
46-3-185, 1-10-025-10-1
558700 2829, 2866, 2867
FILE # 003011.177
W.C.# 27501

Uploaded & Validated by:
Sam H. Pang
03/02/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2466 TO 2472

2466

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/01/99 17:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0013	.1527	.0022	22.40
Stddev	.0004	.0003	.0029	.0004	.31
%RSD	272.9	26.69	1.868	16.96	1.391
#1	-.0004	.0016	.1517	.0019	21.93
#2	.0002	.0016	.1551	.0026	22.53
#3	.0001	.0013	.1563	.0023	22.64
#4	.0000	.0010	.1503	.0018	22.65
#5	.0008	.0009	.1500	.0025	22.22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.680
Stddev	.107
%RSD	.84525

#1	12.798
#2	12.662
#3	12.575
#4	12.580
#5	12.784

Method: ISOU

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 17:14 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	.0002	-.0002	.0002	-.0002
Stddev	.0003	.0003	.0002	.0001	.0002
%RSD	483.2	128.1	107.2	71.99	94.47

#1	-.0001	-.0001	-.0001	.0002	-.0004
#2	-.0004	.0003	-.0001	.0004	.0001
#3	-.0003	.0000	-.0005	.0001	-.0001
#4	.0001	.0003	-.0002	.0000	-.0003
#5	.0004	.0006	.0000	.0002	-.0002

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.772
Stddev	.267
%RSD	2.0932

#1	13.109
#2	12.428
#3	12.626
#4	12.951
#5	12.747

Analysis Report

03/01/99 17:22:32

page 1

Method: ISOU Sample Name: s98t002829 Operator: bjg
 Comment: 40501 df
 Run Time: 03/01/99 17:21 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	-.0002	-.0001	.0000	.0113
Stddev	.0004	.0006	.0005	.0005	.0017
%RSD	228.2	253.2	875.9	16910.	15.10

#1	-.0008	-.0009	-.0007	-.0006	.0083
#2	-.0002	.0003	.0000	.0002	.0117
#3	.0002	-.0007	-.0004	.0008	.0116
#4	.0002	-.0002	.0004	-.0002	.0121
#5	-.0003	.0003	.0004	-.0002	.0126

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.370
Stddev	1.630
%RSD	12.194

#1	16.270
#2	12.703
#3	12.653
#4	12.849
#5	12.373

Analysis Report

03/01/99 17:25:12

page 1

Method: ISOU Sample Name: s98t002829_d Operator: bjg
 Comment: 40501 df
 Run Time: 03/01/99 17:23 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0002	-.0003	-.0003	.0114
Stddev	.0004	.0001	.0002	.0003	.0011
%RSD	142.9	50.39	62.67	96.34	9.599

#1	-.0007	-.0003	-.0001	-.0006	.0112
#2	.0001	-.0002	-.0001	.0000	.0099
#3	-.0005	-.0002	-.0002	.0000	.0114
#4	.0000	-.0001	-.0005	-.0008	.0114
#5	-.0002	-.0003	-.0004	-.0003	.0130

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.709
Stddev	.418
%RSD	3.2870

#1	12.946
#2	12.157
#3	13.135
#4	12.926
#5	12.378

2470

Analysis Report

03/01/99 17:27:22

page 1

Method: ISOU Sample Name: s98t002966 Operator: bjg
 Comment: 40501 df
 Run Time: 03/01/99 17:26 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0002	-.0003	-.0004	.0135
Stddev	.0003	.0001	.0005	.0002	.0007
%RSD	97.10	56.30	197.7	50.69	4.848

#1	-.0002	-.0001	-.0009	-.0004	.0140
#2	-.0009	-.0003	-.0002	-.0001	.0138
#3	-.0001	-.0001	.0003	-.0006	.0135
#4	.0000	-.0001	.0001	-.0005	.0123
#5	-.0005	-.0003	-.0006	-.0003	.0137

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.337
Stddev	.294
%RSD	2.2034

#1	13.059
#2	13.292
#3	13.172
#4	13.337
#5	13.825

2471

Analysis Report

03/01/99 17:30:13

page 1

Method: ISOU

Sample Name: s98t002966_d

Operator: bjg

Comment: 40501 df

Run Time: 03/01/99 17:28 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	-.0006	-.0006	-.0001	.0134
Stddev	.0002	.0003	.0002	.0002	.0008
%RSD	28.05	56.76	42.42	268.8	5.896
#1	-.0004	-.0004	-.0009	.0001	.0132
#2	-.0008	-.0005	-.0003	-.0004	.0127
#3	-.0004	-.0010	-.0007	-.0002	.0132
#4	-.0007	-.0002	-.0004	.0002	.0131
#5	-.0007	-.0009	-.0006	-.0002	.0148

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.486
Stddev	.103
%RSD	.76650

#1	13.598
#2	13.355
#3	13.455
#4	13.436
#5	13.584

2472

Analysis Report

03/01/99 17:32:44

page 1

Method: ISOU   Sample Name: s98t002967   Operator: bjg
 Comment: 40501 df
 Run Time: 03/01/99 17:31 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0003	-.0006	-.0002	.0129
Stddev	.0002	.0004	.0004	.0004	.0010
%RSD	59.30	134.3	68.87	265.6	8.057

#1	-.0006	-.0002	-.0008	.0001	.0123
#2	-.0004	.0003	-.0011	-.0006	.0147
#3	-.0002	-.0007	-.0006	.0005	.0122
#4	-.0004	-.0003	-.0004	-.0003	.0132
#5	-.0001	-.0006	.0000	-.0004	.0123

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.685
Stddev	.194
%RSD	1.4170

#1	13.699
#2	13.496
#3	13.896
#4	13.854
#5	13.481

2473

Analysis Report

03/01/99 17:35:14

page 1

Method: ISOU

Sample Name: s98t002967_d

Operator: bjg

Comment: 40501 df

Run Time: 03/01/99 17:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	-.0003	-.0003	-.0003	.0141
Stddev	.0003	.0004	.0002	.0005	.0006
%RSD	57.88	121.6	67.94	155.3	4.256
#1	-.0002	-.0004	-.0002	.0000	.0147
#2	-.0006	.0001	-.0001	-.0009	.0132
#3	-.0011	-.0009	-.0002	-.0001	.0146
#4	-.0005	-.0001	-.0005	.0002	.0139
#5	-.0004	-.0002	-.0006	-.0008	.0142

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.921
Stddev	.149
%RSD	1.0726

#1	13.700
#2	13.960
#3	13.887
#4	14.113
#5	13.945

2474

Analysis Report

03/01/99 17:37:14

page 1

Method: ISOU

Sample Name: s98t002967_a

Operator: bjg

Comment: 40501 df

Run Time: 03/01/99 17:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	.0009	.1446	.0020	21.13
Stddev	.0005	.0004	.0026	.0005	.54
%RSD	109.3	45.40	1.806	22.87	2.546
#1	-.0006	.0006	.1405	.0020	20.32
#2	-.0004	.0008	.1466	.0024	21.79
#3	.0002	.0015	.1462	.0014	21.12
#4	-.0011	.0005	.1436	.0024	21.07
#5	-.0002	.0009	.1463	.0016	21.37

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.149
Stddev	.359
%RSD	2.5369

#1	14.750
#2	13.866
#3	14.010
#4	14.198
#5	13.920

2475

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/01/99 17:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	.0012	.1570	.0019	22.66
Stddev	.0003	.0003	.0028	.0002	.22
%RSD	130.6	28.58	1.762	9.262	.9857
#1	-.0001	.0010	.1575	.0019	22.33
#2	-.0006	.0011	.1583	.0016	22.64
#3	.0000	.0016	.1576	.0020	22.96
#4	.0001	.0016	.1594	.0019	22.64
#5	-.0004	.0008	.1523	.0020	22.71

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.009
Stddev	.091
%RSD	.65084

#1	14.021
#2	13.982
#3	13.871
#4	14.055
#5	14.115

Analysis Report

03/01/99 17:45:46

page.1

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 17:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	-.0007	-.0006	-.0004	-.0006
Stddev	.0004	.0005	.0003	.0001	.0003
%RSD	73.32	71.95	46.99	26.53	58.23

#1	-.0002	-.0008	-.0002	-.0002	-.0002
#2	-.0011	-.0013	-.0010	-.0005	-.0003
#3	-.0005	.0000	-.0006	-.0005	-.0008
#4	-.0009	-.0004	-.0008	-.0005	-.0008
#5	-.0001	-.0007	-.0005	-.0005	-.0008

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.396
Stddev	.190
%RSD	1.3177

#1	14.496
#2	14.363
#3	14.515
#4	14.528
#5	14.078

2477

LABCORE Completed Worklist Report for Worklist# 27502

Analyst: bjg

Instrument: ICPMS1

Book#: 425B 42B57, 990121-1
SO 1/21/99

Method: LA-506-101

Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSPR-QC TC99	QC	0.0197	1.60e-02	81.218 % Recovery	
1	ICV	0	QMSPR-QC SN117	QC	0.0077	6.46e-03	83.896 % Recovery	
1	ICV	0	QMSPR-QC SN126	QC	n/a	<1.50e-4	% Recovery	
1	ICV	0	QMSPR-QC CS133	QC	0.020	2.14e-02	107.000 % Recovery	
1	ICV	0	QMSPR-QC CS135	QC	n/a	3.88e-05	% Recovery	
1	ICV	0	QMSPR-QC CS137	QC	n/a	6.41e-05	% Recovery	
2	ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC CS133	QC	1	6.90e-06	6.900e-006 ug/mL	
2	ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 TC99-D1	LIQUID	N/A	5.464e+00	8.10e-002 ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 SN117-D1	LIQUID	N/A <	6.075e+00	6.075 ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 SN126-D1	LIQUID	N/A <	6.075e+00	6.075 ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 CS133-D1	LIQUID	N/A	8.513e+00	8.10e-002 ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 CS135-D1	LIQUID	N/A	2.616e+00	8.10e-002 ug/mL	
3	SAMPLE	S98T002968	0 D QMSPR-D1 CS137-D1	LIQUID	N/A	2.341e+00	8.10e-002 ug/mL	
4	DUP	S98T002968	0 D QMSPR-D1 TC99-D1	LIQUID	5.46e+00	5.42e+00	0.735 RPD	
4	DUP	S98T002968	0 D QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0	RPD	
4	DUP	S98T002968	0 D QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0	RPD	
4	DUP	S98T002968	0 D QMSPR-D1 CS133-D1	LIQUID	8.51e+00	8.33e+00	2.138 RPD	
4	DUP	S98T002968	0 D QMSPR-D1 CS135-D1	LIQUID	2.62e+00	2.49e+00	5.088 RPD	
4	DUP	S98T002968	0 D QMSPR-D1 CS137-D1	LIQUID	2.34e+00	2.32e+00	0.858 RPD	
5	SPK-POST	S98T002968	0 D QMSPR-D1 TC99-D1	LIQUID	0.0219	1.81e-02	82.648 % Recovery	
5	SPK-POST	S98T002968	0 D QMSPR-D1 SN117-D1	LIQUID	n/a	4.36e-03	% Recovery	
5	SPK-POST	S98T002968	0 D QMSPR-D1 SN126-D1	LIQUID	n/a	8.00e-07	% Recovery	
5	SPK-POST	S98T002968	0 D QMSPR-D1 CS133-D1	LIQUID	0.0201	2.16e-02	107.463 % Recovery	
5	SPK-POST	S98T002968	0 D QMSPR-D1 CS135-D1	LIQUID	n/a	-9.00e-7	% Recovery	
5	SPK-POST	S98T002968	0 D QMSPR-D1 CS137-D1	LIQUID	n/a	7.00e-07	% Recovery	
6	SAMPLE	S98T002969	0 D QMSPR-D1 TC99-D1	LIQUID	N/A	5.512e+00	8.10e-002 ug/mL	
6	SAMPLE	S98T002969	0 D QMSPR-D1 SN117-D1	LIQUID	N/A <	6.075e+00	6.075 ug/mL	
6	SAMPLE	S98T002969	0 D QMSPR-D1 SN126-D1	LIQUID	N/A <	6.075e+00	6.075 ug/mL	
6	SAMPLE	S98T002969	0 D QMSPR-D1 CS133-D1	LIQUID	N/A	8.015e+00	8.10e-002 ug/mL	
6	SAMPLE	S98T002969	0 D QMSPR-D1 CS135-D1	LIQUID	N/A	2.406e+00	8.10e-002 ug/mL	
6	SAMPLE	S98T002969	0 D QMSPR-D1 CS137-D1	LIQUID	N/A	2.232e+00	8.10e-002 ug/mL	
7	DUP	S98T002969	0 D QMSPR-D1 TC99-D1	LIQUID	5.51e+00	5.45e+00	1.095 RPD	
7	DUP	S98T002969	0 D QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0	RPD	
7	DUP	S98T002969	0 D QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0	RPD	
7	DUP	S98T002969	0 D QMSPR-D1 CS133-D1	LIQUID	8.02e+00	7.82e+00	2.525 RPD	
7	DUP	S98T002969	0 D QMSPR-D1 CS135-D1	LIQUID	2.41e+00	2.51e+00	4.065 RPD	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2478

LABCORE Completed Worklist Report for Worklist# 27502

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 DUP	S98T002969	0 D	QMSPR-D1 CS137-D1	LIQUID	2.23e+00	2.28e+00	2.217	RPD
8 SAMPLE	S98T003010	0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	4.832e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003010	0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
8 SAMPLE	S98T003010	0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
8 SAMPLE	S98T003010	0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	7.250e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003010	0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.268e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003010	0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	1.956e+00	8.10e-002	ug/mL
9 DUP	S98T003010	0 D	QMSPR-D1 TC99-D1	LIQUID	4.83e+00	4.83e+00	0.000	RPD
9 DUP	S98T003010	0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
9 DUP	S98T003010	0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
9 DUP	S98T003010	0 D	QMSPR-D1 CS133-D1	LIQUID	7.25e+00	6.93e+00	4.513	RPD
9 DUP	S98T003010	0 D	QMSPR-D1 CS135-D1	LIQUID	2.27e+00	2.11e+00	7.306	RPD
9 DUP	S98T003010	0 D	QMSPR-D1 CS137-D1	LIQUID	1.96e+00	1.92e+00	2.062	RPD
10 CCV	0		QMSPR-QC TC99	QC	0.0197	1.58e-02	80.203 %	Recovery
10 CCV	0		QMSPR-QC SN117	QC	0.0077	6.70e-03	87.013 %	Recovery
10 CCV	0		QMSPR-QC SN126	QC	n/a	<1.50e-4		% Recovery
10 CCV	0		QMSPR-QC CS133	QC	0.020	2.10e-02	105.000 %	Recovery
10 CCV	0		QMSPR-QC CS135	QC	n/a	3.62e-05		% Recovery
10 CCV	0		QMSPR-QC CS137	QC	n/a	6.59e-05		% Recovery
11 CCB	0		QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
11 CCB	0		QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
11 CCB	0		QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27502

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. Pang 1/21/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27502

Analyst: BIG Instrument: ICPMS1 Book# 42857 980121-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC	QC		
2 ICB			@MSPR-QC	QC		
3 SAMPLE	S98T002968	0 D	@MSPR-D1	LIQUID	98000509	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
4 DUP	S98T002968	0 D	@MSPR-D1	LIQUID		
5 SPK-POST	S98T002968	0 D	@MSPR-D1	LIQUID		
6 SAMPLE	S98T002969	0 D	@MSPR-D1	LIQUID	98000509	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
7 DUP	S98T002969	0 D	@MSPR-D1	LIQUID		
8 SAMPLE	S98T003010	0 D	@MSPR-D1	LIQUID	98000510	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
9 DUP	S98T003010	0 D	@MSPR-D1	LIQUID		
10 CCV			@MSPR-QC	QC		
11 CCB			@MSPR-QC	QC		

Final page for worklist # 27502

Analyst Signature BIG Date 1-21-99

Analyst Signature _____ Date _____

S98T002968, 2969, 3010
-1-10ml - .025-10ml 2F
46501

Data Entry Comments:

SCANS = BACKGROUND BLANK (BLACK), 982968, 982969, 983010

Uploaded + Validated by:

Sant M. Pang

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1/21/99

Analysis Report

01/21/99 13:33:41

page 1

Method: MS_TUNE Sample Name: ms-tune

Operator: bjs

Comment:

Run Time: 01/21/99 13:32 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	100200.
Stddev	1083.
%RSD	1.081

#1	100800.
#2	99340.
#3	98270.
#4	100300.
#5	100400.
#6	100100.
#7	99410.
#8	100200.
#9	101500.
#10	102000.

1-21-99

64-105

1-10-10-025-10-1

588700 2968, 65, 3010

FILED PL 0121A.TFF

W.L.B. 27802

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2481 TO 2498.

Analysis Report

01/21/99 13:36:37

page 1

Method: TC_SN_CS Sample Name: Blank

Operator:

Comment:

Run Time: 01/21/99 13:34 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.074	2.017	10.08	2.514	1.393
Stddev	.353	.175	.37	.476	.216
%RSD	17.04	8.693	3.719	18.94	15.50
#1	1.771	2.066	10.48	2.262	1.131
#2	2.566	2.082	9.685	3.147	1.307
#3	2.258	1.865	10.06	2.798	1.620
#4	2.060	1.820	10.44	1.916	1.293
#5	1.713	2.251	9.738	2.447	1.615

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.596	9.004	13.59	79.94
Stddev	.891	.660	1.18	2.69
%RSD	13.51	7.332	8.688	3.364
#1	6.197	9.492	15.00	82.18
#2	7.941	7.941	12.59	76.70
#3	6.578	9.229	12.96	82.82
#4	5.509	8.814	12.65	77.70
#5	6.753	9.542	14.73	80.30

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.267
Stddev	.114
%RSD	1.1071

#1	10.167
#2	10.326
#3	10.185
#4	10.437
#5	10.218

2482

Method: TC_SN_CS Sample Name: PRIV Std

Operator:

Comment:

Run Time: 01/21/99 13:38 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]
Line	99/pulse	100/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	13730.	13860.	7798.	7798.
Stddev	260.	152.	100.	100.
%RSD	1.894	1.097	1.285	1.285

#1	13830.	13870.	7692.	7692.
#2	13840.	13980.	7854.	7854.
#3	13490.	13700.	7770.	7770.
#4	13450.	13720.	7942.	7942.
#5	14060.	14030.	7733.	7733.

Elem	Te[125]	Cs[133]	Cs[135]	Cs[137]
Line	125/pulse	133/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1653.	89010.	6374.	11060.
Stddev	10.	1402.	69.	178.
%RSD	.5833	1.575	1.085	1.605

#1	1646.	88010.	6338.	11020.
#2	1664.	90330.	6411.	11120.
#3	1641.	87100.	6356.	10930.
#4	1659.	90100.	6294.	10890.
#5	1654.	89530.	6473.	11330.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	67080.
Stddev	612.
%RSD	.9123

#1	66370.
#2	67740.
#3	66490.
#4	67340.
#5	67470.

Sample Name: PRIV Std Run Time: 01/21/99 13:38

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.3759
Stddev	.1196
%RSD	1.2753

#1	9.4917
#2	9.3243
#3	9.4994
#4	9.3460
#5	9.2179

2484

Analysis Report

Method: TC_SN_CS Sample Name: blank

Comment: background

Operator: bjg

Run Time: 01/21/99 13:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	.0000	.0009	-.0001	-.0014
Stddev	.0004	.0002	.0007	.0003	.0006
%RSD	649.4	574.3	77.92	274.3	42.41

#1	-.0001	.0002	-.0001	-.0001	-.0013
#2	-.0004	.0002	.0005	-.0003	-.0020
#3	.0006	-.0004	.0014	.0003	-.0005
#4	-.0002	.0000	.0013	.0001	-.0020
#5	-.0001	.0002	.0013	-.0005	-.0015

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0115	.0000	.0002	.0004
Stddev	.0004	.001	.0010	.0002
%RSD	3.270	5562.	406.0	56.65

#1	.0110	-.0003	-.0014	.0007
#2	.0120	.0007	.0001	.0003
#3	.0112	-.0003	.0010	.0004
#4	.0116	-.0005	.0010	.0001
#5	.0115	.0003	.0006	.0004

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.128
Stddev	.019
%RSD	.19077

#1	10.148
#2	10.149
#3	10.116
#4	10.115
#5	10.110

Method: TC_SN_CS   Sample Name: icv   Operator: bjg

Comment: msprv. lcs

Run Time: 01/21/99 13:54   Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.01	.0069	1.336	.0001	-.0007
Stddev	.22	.0005	.050	.0003	.0008
%RSD	1.397	7.342	3.726	210.3	112.4
#1	15.73	.0076	1.289	-.0001	-.0018
#2	16.24	.0070	1.288	.0003	-.0008
#3	16.15	.0067	1.346	-.0003	-.0008
#4	15.82	.0068	1.351	.0005	-.0008
#5	16.10	.0062	1.407	.0002	.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.39	.0388	.0641	.0453
Stddev	.23	.0022	.0022	.0030
%RSD	1.084	5.699	3.492	6.583
#1	20.99	.0376	.0647	.0422
#2	21.49	.0420	.0668	.0451
#3	21.53	.0368	.0638	.0443
#4	21.55	.0375	.0606	.0447
#5	21.41	.0403	.0648	.0502

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.9853
Stddev	.1332
%RSD	1.3340
#1	10.215
#2	9.8875
#3	9.9717
#4	9.9549
#5	9.8979

Method: TC_SN_CS Sample Name: icv

Operator: bjj

Comment: 7.7 ppb Sn

Run Time: 01/21/99 14:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0059	-.0003	6.459	.0004	-.0014
Stddev	.0004	.0002	.060	.0007	.0013
%RSD	7.143	53.36	.9278	154.2	87.66

#1	.0065	-.0005	6.386	.0011	-.0013
#2	.0054	-.0001	6.478	.0000	.0006
#3	.0058	-.0002	6.422	-.0005	-.0021
#4	.0057	-.0004	6.465	.0008	-.0016
#5	.0063	-.0004	6.544	.0008	-.0027

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0127	-.0049	-.0095	-.0071
Stddev	.0024	.0005	.0006	.0001
%RSD	19.14	11.05	6.278	1.561

#1	.0162	-.0053	-.0104	-.0071
#2	.0133	-.0055	-.0097	-.0072
#3	.0095	-.0047	-.0088	-.0072
#4	.0119	-.0041	-.0093	-.0073
#5	.0125	-.0050	-.0094	-.0070

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4204
Stddev	.1652
%RSD	1.7539

#1	9.6533
#2	9.4479
#3	9.4217
#4	9.3892
#5	9.1897

Method: TC_SN_CS Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/21/99 14:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0000	.0008	.0002	-.0017
Stddev	.0002	.0000	.0003	.0003	.0011
%RSD	194.6	900.1	39.79	177.3	65.01

#1	.0002	-.0001	.0011	.0000	-.0018
#2	-.0001	-.0004	.0006	.0004	-.0027
#3	-.0001	-.0003	.0011	-.0002	-.0030
#4	.0003	.0003	.0004	.0005	-.0008
#5	.0002	.0003	.0006	.0001	-.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0069	.0000	.0004	.0001
Stddev	.0007	.0009	.0007	.0006
%RSD	9.677	7910.	188.7	447.6

#1	.0066	.0005	.0012	.0011
#2	.0077	-.0011	-.0001	-.0002
#3	.0073	.0006	.0009	-.0002
#4	.0068	.0007	.0005	.0001
#5	.0060	-.0007	-.0006	-.0002

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8059
Stddev	.3237
%RSD	3.3009

#1	9.2420
#2	9.9117
#3	9.9316
#4	9.8752
#5	10.069

2

Analysis Report

Method: TC_SN_CS Sample Name: s98t002968 Operator: bjg
 Comment: 40501 df
 Run Time: 01/21/99 14:15 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1349	.1671	.0405	.0010	-.0009
Stddev	.0050	.0101	.0029	.0003	.0007
%RSD	3.716	6.053	7.232	32.46	82.30

#1	.1432	.1839	.0455	.0009	-.0006
#2	.1303	.1653	.0403	.0013	-.0004
#3	.1353	.1648	.0395	.0006	-.0003
#4	.1331	.1565	.0381	.0014	-.0021
#5	.1324	.1647	.0389	.0008	-.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2102	.0646	.0578	-.0073
Stddev	.0081	.0044	.0046	.0002
%RSD	3.836	6.874	8.040	2.276

#1	.2230	.0723	.0648	-.0070
#2	.2046	.0620	.0559	-.0074
#3	.2064	.0641	.0586	-.0075
#4	.2038	.0623	.0520	-.0073
#5	.2130	.0620	.0575	-.0072

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.7324
Stddev	.3301
%RSD	3.3919

#1	9.1651
#2	9.9958
#3	9.7872
#4	9.9299
#5	9.7842

Method: TC_SN_CS Sample Name: s98t002968_d

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1339	.1651	.0401	.0012	-.0017
Stddev	.0038	.0031	.0015	.0003	.0014
%RSD	2.871	1.876	3.762	24.87	86.51

#1	.1350	.1693	.0404	.0010	-.0020
#2	.1328	.1628	.0376	.0009	.0001
#3	.1400	.1675	.0414	.0012	-.0014
#4	.1302	.1628	.0411	.0013	-.0012
#5	.1314	.1629	.0399	.0016	-.0038

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.2057	.0614	.0574	-.0076
Stddev	.0045	.0016	.0027	.0001
%RSD	2.205	2.581	4.646	.8749

#1	.2115	.0623	.0611	-.0075
#2	.2022	.0603	.0586	-.0076
#3	.2008	.0624	.0539	-.0076
#4	.2091	.0629	.0562	-.0076
#5	.2050	.0592	.0573	-.0077

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8123
Stddev	.0592
%RSD	.60379

#1	9.7303
#2	9.8785
#3	9.8010
#4	9.7907
#5	9.8607

Method: TC_SN_CS Sample Name: s98t002968_a

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.23	.1657	4.362	.0008	-.0002
Stddev	.37	.0020	.065	.0004	.0008
%RSD	2.033	1.183	1.494	42.76	348.3
#1	18.22	.1643	4.353	.0009	-.0001
#2	18.87	.1676	4.446	.0005	-.0010
#3	18.09	.1681	4.410	.0014	-.0005
#4	18.01	.1639	4.297	.0008	.0010
#5	17.97	.1647	4.305	.0006	-.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.85	.0637	.0585	-.0063
Stddev	.41	.0017	.0014	.0001
%RSD	1.860	2.694	2.329	1.243
#1	21.54	.0649	.0593	-.0063
#2	22.54	.0653	.0562	-.0062
#3	21.85	.0609	.0597	-.0063
#4	21.73	.0638	.0590	-.0064
#5	21.57	.0635	.0586	-.0063

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4481
Stddev	.1821
%RSD	1.9271

#1	9.5306
#2	9.1284
#3	9.4775
#4	9.5280
#5	9.5760

Method: TC_SN_CS Sample Name: s98t002969

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1361	.1632	.0379	.0010	-.0017
Stddev	.0016	.0046	.0020	.0004	.0007
%RSD	1.183	2.799	5.406	38.28	45.13

#1	.1351	.1671	.0372	.0007	-.0025
#2	.1387	.1651	.0354	.0015	-.0008
#3	.1357	.1587	.0379	.0008	-.0015
#4	.1365	.1579	.0378	.0007	-.0024
#5	.1346	.1673	.0411	.0013	-.0012

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1979	.0594	.0551	-.0073
Stddev	.0070	.0025	.0012	.0001
%RSD	3.546	4.156	2.154	1.533

#1	.1934	.0568	.0570	-.0075
#2	.2004	.0593	.0551	-.0074
#3	.1923	.0573	.0549	-.0072
#4	.1944	.0607	.0539	-.0072
#5	.2091	.0628	.0544	-.0074

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.5349
Stddev	.1649
%RSD	1.7295

#1	9.7926
#2	9.4719
#3	9.5736
#4	9.4875
#5	9.3487

Q.

Method: TC_SN_CS Sample Name: s98t002969_d

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1345	.1659	.0372	.0014	-.0009
Stddev	.0042	.0030	.0019	.0003	.0008
%RSD	3.086	1.813	5.212	22.99	82.91
#1	.1327	.1617	.0345	.0015	-.0009
#2	.1301	.1639	.0358	.0011	.0000
#3	.1321	.1677	.0379	.0012	-.0015
#4	.1397	.1677	.0392	.0019	-.0019
#5	.1381	.1687	.0384	.0014	-.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1932	.0620	.0562	-.0074
Stddev	.0031	.0024	.0012	.0000
%RSD	1.589	3.804	2.056	.6529
#1	.1919	.0599	.0553	-.0074
#2	.1893	.0607	.0555	-.0074
#3	.1929	.0602	.0556	-.0074
#4	.1976	.0650	.0563	-.0073
#5	.1942	.0640	.0581	-.0073

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.6263
Stddev	.0932
%RSD	.96860

#1	9.7230
#2	9.6771
#3	9.6368
#4	9.4757
#5	9.6188

Method: TC_SN_CS Sample Name: s98t003010

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1193	.1512	.0399	.0011	-.0016
Stddev	.0137	.0157	.0057	.0004	.0009
%RSD	11.52	10.40	14.19	32.30	55.22
#1	.1069	.1383	.0349	.0013	-.0016
#2	.1176	.1500	.0372	.0011	-.0017
#3	.1146	.1443	.0380	.0013	-.0015
#4	.1147	.1453	.0399	.0005	-.0028
#5	.1429	.1784	.0495	.0015	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1790	.0560	.0483	-.0077
Stddev	.0191	.0075	.0058	.0001
%RSD	10.65	13.36	11.94	.8754
#1	.1669	.0493	.0452	-.0077
#2	.1699	.0543	.0431	-.0077
#3	.1712	.0554	.0476	-.0077
#4	.1742	.0523	.0474	-.0077
#5	.2127	.0688	.0580	-.0076

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.3041
Stddev	.7515
%RSD	8.0769

#1	9.8927
#2	9.5475
#3	9.5600
#4	9.5336
#5	7.9869

Method: TC_SN_CS Sample Name: s98t003010_d

Operator: bjg

Comment: 40501 df

Run Time: 01/21/99 14:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1192	.1462	.0345	.0010	-.0013
Stddev	.0026	.0018	.0016	.0002	.0013
%RSD	2.198	1.231	4.640	20.33	102.5
#1	.1185	.1481	.0347	.0010	-.0022
#2	.1157	.1472	.0335	.0012	-.0015
#3	.1214	.1461	.0365	.0009	.0006
#4	.1182	.1461	.0324	.0010	-.0026
#5	.1222	.1433	.0353	.0007	-.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1712	.0522	.0473	-.0077
Stddev	.0036	.0015	.0013	.0001
%RSD	2.091	2.870	2.744	1.214
#1	.1659	.0523	.0481	-.0078
#2	.1709	.0531	.0486	-.0076
#3	.1757	.0534	.0476	-.0076
#4	.1707	.0527	.0469	-.0078
#5	.1729	.0497	.0453	-.0077

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.5533
Stddev	.0578
%RSD	.60529

#1	9.6088
#2	9.6035
#3	9.4695
#4	9.5583
#5	9.5262

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/21/99 15:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0061	-.0004	6.695	.0002	-.0004
Stddev	.0004	.0003	.433	.0004	.0013
%RSD	6.435	93.85	6.468	220.8	356.9

#1	.0059	-.0002	7.462	.0007	.0008
#2	.0066	.0001	6.459	.0002	-.0021
#3	.0056	-.0007	6.462	-.0001	.0004
#4	.0063	-.0006	6.603	-.0002	.0004
#5	.0061	-.0004	6.488	.0003	-.0013

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0014	-.0053	-.0094	-.0072
Stddev	.0013	.0003	.0004	.0001
%RSD	90.85	5.286	4.318	.8543

#1	.0033	-.0051	-.0094	-.0071
#2	.0008	-.0049	-.0095	-.0072
#3	.0017	-.0055	-.0091	-.0072
#4	.0011	-.0053	-.0101	-.0072
#5	-.0001	-.0056	-.0091	-.0072

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.5914
Stddev	.4058
%RSD	4.7236

#1	7.8679
#2	8.7983
#3	8.7159
#4	8.7789
#5	8.7961

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/21/99 15:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	15.80	.0071	1.413	.0000	-.0018
Stddev	.36	.0004	.120	.0004	.0007
%RSD	2.308	5.123	8.506	1996.	36.31
#1	15.25	.0076	1.242	-.0005	-.0009
#2	15.96	.0070	1.509	.0002	-.0022
#3	16.24	.0067	1.544	.0005	-.0024
#4	15.85	.0072	1.366	-.0003	-.0013
#5	15.71	.0069	1.405	.0001	-.0023

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	20.99	.0362	.0659	.0434
Stddev	.42	.0017	.0026	.0015
%RSD	2.002	4.557	3.964	3.425
#1	20.28	.0337	.0617	.0414
#2	21.23	.0378	.0675	.0446
#3	20.93	.0365	.0686	.0446
#4	21.34	.0357	.0660	.0441
#5	21.15	.0375	.0658	.0422

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1686
Stddev	.1459
%RSD	1.5908

#1	9.3814
#2	9.0573
#3	9.0296
#4	9.1264
#5	9.2481

Method: TC_SN_CS Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/21/99 15:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	-.0002	.0003	-.0001	-.0015
Stddev	.0007	.0001	.0007	.0002	.0010
%RSD	106.8	47.40	214.0	269.4	67.18

#1	.0017	-.0003	.0000	-.0003	-.0011
#2	.0004	-.0002	-.0001	.0002	-.0005
#3	.0003	-.0001	.0012	-.0002	-.0011
#4	.0008	-.0001	.0009	.0002	-.0015
#5	-.0001	-.0004	-.0004	-.0003	-.0031

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0016	.0003	.0012	.0002
Stddev	.0013	.0011	.0013	.0010
%RSD	81.95	345.4	108.7	644.5

#1	.0009	.0000	.0027	-.0003
#2	.0040	.0013	.0016	.0019
#3	.0015	.0014	.0016	.0000
#4	.0011	.0002	.0006	.0001
#5	.0008	-.0013	-.0007	-.0008

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1609
Stddev	.9037
%RSD	9.8651

#1	9.2411
#2	7.8709
#3	9.1614
#4	9.1080
#5	10.423

LABCORE Completed Worklist Report for Worklist# 27503

Analyst: bjg

Instrument: ICPMS1

Book#: 990104-1

Method: LA-506-101

Rev/Mod A-0

Worklist Comment: 1CP/MS BY-105 (DIRECT) MSU

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	QMSU-QC U235	QC	0.000143	1.45e-04	101.399 % Recovery	
1 ICV	0	QMSU-QC U238	QC	0.020	2.12e-02	106.000 % Recovery	
2 ICB	0	QMSU-QC U233	QC	1	<3.23e-5		ug/mL
2 ICB	0	QMSU-QC U234	QC	1	<4.44e-5		ug/mL
2 ICB	0	QMSU-QC U235	QC	1	<1.64e-5		ug/mL
2 ICB	0	QMSU-QC U236	QC	1	<2.27e-5		ug/mL
2 ICB	0	QMSU-QC U238	QC	1	<2.14e-5		ug/mL
3 SAMPLE	S98T002968 0 D	QMSU-D1 U233-D1	LIQUID	N/A <	1.307e+00	1.307	ug/mL
3 SAMPLE	S98T002968 0 D	QMSU-D1 U234-D1	LIQUID	N/A <	1.800e+00	1.800	ug/mL
3 SAMPLE	S98T002968 0 D	QMSU-D1 U235-D1	LIQUID	N/A <	6.647e-01	0.665	ug/mL
3 SAMPLE	S98T002968 0 D	QMSU-D1 U236-D1	LIQUID	N/A <	9.210e-01	0.921	ug/mL
3 SAMPLE	S98T002968 0 D	QMSU-D1 U238-D1	LIQUID	N/A <	8.647e-01	0.865	ug/mL
4 DUP	S98T002968 0 D	QMSU-D1 U233-D1	LIQUID	<1.31e0	<1.31e0		RPD
4 DUP	S98T002968 0 D	QMSU-D1 U234-D1	LIQUID	<1.80e0	<1.80e0		RPD
4 DUP	S98T002968 0 D	QMSU-D1 U235-D1	LIQUID	<6.65e-1	<6.65e-1		RPD
4 DUP	S98T002968 0 D	QMSU-D1 U236-D1	LIQUID	<9.21e-1	<9.21e-1		RPD
4 DUP	S98T002968 0 D	QMSU-D1 U238-D1	LIQUID	<8.65e-1	<8.65e-1		RPD
5 SPK-POST	S98T002968 0 D	QMSU-D1 U233-D1	LIQUID	n/a	0.00e+00		% Recovery
5 SPK-POST	S98T002968 0 D	QMSU-D1 U234-D1	LIQUID	n/a	0.00e+00		% Recovery
5 SPK-POST	S98T002968 0 D	QMSU-D1 U235-D1	LIQUID	n/a	1.31e-04		% Recovery
5 SPK-POST	S98T002968 0 D	QMSU-D1 U236-D1	LIQUID	n/a	-0.00e0		% Recovery
5 SPK-POST	S98T002968 0 D	QMSU-D1 U238-D1	LIQUID	0.0200	1.95e-02	97.500	% Recovery
6 SAMPLE	S98T002969 0 D	QMSU-D1 U233-D1	LIQUID	N/A <	1.307e+00	1.307	ug/mL
6 SAMPLE	S98T002969 0 D	QMSU-D1 U234-D1	LIQUID	N/A <	1.800e+00	1.800	ug/mL
6 SAMPLE	S98T002969 0 D	QMSU-D1 U235-D1	LIQUID	N/A <	6.647e-01	0.665	ug/mL
6 SAMPLE	S98T002969 0 D	QMSU-D1 U236-D1	LIQUID	N/A <	9.210e-01	0.921	ug/mL
6 SAMPLE	S98T002969 0 D	QMSU-D1 U238-D1	LIQUID	N/A <	8.647e-01	0.865	ug/mL
7 DUP	S98T002969 0 D	QMSU-D1 U233-D1	LIQUID	<1.31e0	<1.31e0		RPD
7 DUP	S98T002969 0 D	QMSU-D1 U234-D1	LIQUID	<1.80e0	<1.80e0		RPD
7 DUP	S98T002969 0 D	QMSU-D1 U235-D1	LIQUID	<6.65e-1	<6.65e-1		RPD
7 DUP	S98T002969 0 D	QMSU-D1 U236-D1	LIQUID	<9.21e-1	<9.21e-1		RPD
7 DUP	S98T002969 0 D	QMSU-D1 U238-D1	LIQUID	<8.65e-1	<8.65e-1		RPD
8 SAMPLE	S98T003010 0 D	QMSU-D1 U233-D1	LIQUID	N/A <	1.307e+00	1.307	ug/mL
8 SAMPLE	S98T003010 0 D	QMSU-D1 U234-D1	LIQUID	N/A <	1.800e+00	1.800	ug/mL
8 SAMPLE	S98T003010 0 D	QMSU-D1 U235-D1	LIQUID	N/A <	6.647e-01	0.665	ug/mL
8 SAMPLE	S98T003010 0 D	QMSU-D1 U236-D1	LIQUID	N/A <	9.210e-01	0.921	ug/mL
8 SAMPLE	S98T003010 0 D	QMSU-D1 U238-D1	LIQUID	N/A <	8.647e-01	0.865	ug/mL
9 DUP	S98T003010 0 D	QMSU-D1 U233-D1	LIQUID	<1.31e0	<1.31e0		RPD
9 DUP	S98T003010 0 D	QMSU-D1 U234-D1	LIQUID	<1.80e0	<1.80e0		RPD
9 DUP	S98T003010 0 D	QMSU-D1 U235-D1	LIQUID	<6.65e-1	<6.65e-1		RPD
9 DUP	S98T003010 0 D	QMSU-D1 U236-D1	LIQUID	<9.21e-1	<9.21e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27503

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 DUP	S98T003010	0 D	QMSU-D1 U238-D1	LIQUID	<8.65e-1	<8.65e-1		RPD
10 CCV	0		QMSU-QC U235	QC	0.000143	1.49e-04	104.196 % Recovery	
10 CCV	0		QMSU-QC U238	QC	0.020	2.21e-02	110.500 % Recovery	
11 CCB	0		QMSU-QC U233	QC	1	<3.23e-5		ug/mL
11 CCB	0		QMSU-QC U234	QC	1	<4.44e-5		ug/mL
11 CCB	0		QMSU-QC U235	QC	1	<1.64e-5		ug/mL
11 CCB	0		QMSU-QC U236	QC	1	<2.27e-5		ug/mL
11 CCB	0		QMSU-QC U238	QC	1	<2.14e-5		ug/mL

Final page for worklist# 27503

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

1/18/99
_____
Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 27503

Analyst: B. Golett Instrument: ICPMS1 Book# 990104-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	SAMPLE	S98T002968 0 D	@MSU-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1				
4	DUP	S98T002968 0 D	@MSU-D1	LIQUID		
5	SPK-POST	S98T002968 0 D	@MSU-D1	LIQUID		
6	SAMPLE	S98T002969 0 D	@MSU-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1				
7	DUP	S98T002969 0 D	@MSU-D1	LIQUID		
8	SAMPLE	S98T003010 0 D	@MSU-D1	LIQUID	98000510	BY-105 (2)
		Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1				
9	DUP	S98T003010 0 D	@MSU-D1	LIQUID		
10	CCV		@MSU-QC	QC		
11	CCB		@MSU-QC	QC		

Final page for worklist # 27503

B. Golett 1-18-99
Analyst Signature Date

40501
Analyst Signature Date

post spike = .02-1 of
10 ppb USD

Data Entry Comments:

Uploaded & Validated by:

Saul M. Ray

1/18/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/18/99 13:13 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	130900.
Stddev	1241.
%RSD	.9481
#1	131500.
#2	131000.
#3	128600.
#4	132800.
#5	130900.
#6	129200.
#7	132000.
#8	130700.
#9	130800.
#10	131400.

1-18-99

By-101

588TW 2968, 2969, 3010

01-10-025-10

File # U0118A.TFT

W.L. # 27503

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2502 TO 2530

Method: ISO_U

Sample Name: BLANK

Operator:

Comment:

Run Time: 01/18/99

13:17 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem.	U_[235]	U_[238]
Line	235/pulse	238/pulse
Units	Cts/S	Cts/S
Avg	4.624	4.636
Stddev	.588	.402
%RSD	12.71	8.667

#1	5.268	4.201
#2	3.725	5.122
#3	4.406	4.340
#4	4.850	4.985
#5	4.869	4.531

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.4869
Stddev	.0829
%RSD	1.1073

#1	7.4981
#2	7.5168
#3	7.6037
#4	7.4223
#5	7.3938

HNF-1663 REV: 0

Method: ISO_U

Sample Name: U High Std

Operator:

Comment:

Run Time: 01/18/99 13:19 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[235]	U_[238]
Line	235/pulse	238/pulse
Units	Cts/S	Cts/S
Avg	188.2	23210.
Stddev	2.0	471.
%RSD	1.058	2.029
#1	187.1	22420.
#2	188.4	23290.
#3	189.9	23660.
#4	185.5	23460.
#5	190.3	23210.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.4735
Stddev	.2039
%RSD	2.7286

#1	7.8133
#2	7.4719
#3	7.3656
#4	7.4384
#5	7.2786

Method: ISO_U Sample Name: iso_u std

Operator: bjg

Comment: iso_u lcs

Run Time: 01/18/99 13:21 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.675	6.134	203.3	8.735	24250.
Stddev	.550	.786	5.4	1.050	468.
%RSD	11.77	12.82	2.665	12.02	1.930
#1	5.263	5.330	197.3	9.394	24040.
#2	4.306	5.673	199.4	9.227	24080.
#3	4.232	5.967	206.8	7.632	23720.
#4	5.291	7.379	210.7	7.588	24450.
#5	4.283	6.320	202.3	9.831	24940.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.2658
Stddev	.1509
%RSD	2.0765

#1	7.5045
#2	7.3152
#3	7.2064
#4	7.1821
#5	7.1206

HNF-1663 REV. 0

Method: ISO_U Sample Name: iso_u std

Operator: bjg

Comment: iso_u lcs

Run Time: 01/18/99 13:21 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.675	6.134	.7694	8.735	103.7
Stddev	.550	.786	.0210	1.050	2.0
%RSD	11.77	12.82	2.727	12.02	1.930
#1	5.263	5.330	.7463	9.394	102.9
#2	4.306	5.673	.7544	9.227	103.0
#3	4.232	5.967	.7827	7.632	101.5
#4	5.291	7.379	.7979	7.588	104.6
#5	4.283	6.320	.7655	9.831	106.7

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.2658
Stddev	.1509
%RSD	2.0765

#1	7.5045
#2	7.3152
#3	7.2064
#4	7.1821
#5	7.1206

Method: ISO_U Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/18/99 13:30 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.205	3.820	3.726	3.726	3.902
Stddev	.569	.784	.289	.401	.377
%RSD	13.53	20.52	7.767	10.76	9.649
#1	3.655	4.538	3.970	3.214	3.277
#2	4.018	3.953	4.083	3.888	4.147
#3	3.908	3.908	3.656	3.404	4.160
#4	5.132	4.211	3.421	3.947	3.816
#5	4.310	2.492	3.502	4.175	4.108

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.7213
Stddev	.2195
%RSD	2.8425

#1	7.9336
#2	7.7157
#3	7.9330
#4	7.5998
#5	7.4246

Method: ISO_U Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/18/99 13:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.205	3.820	-.0035	3.726	-.0031
Stddev	.569	.784	.0011	.401	.0016
%RSD	13.53	20.52	32.26	10.76	51.28
#1	3.655	4.538	-.0025	3.214	-.0058
#2	4.018	3.953	-.0021	3.888	-.0021
#3	3.908	3.908	-.0037	3.404	-.0020
#4	5.132	4.211	-.0047	3.947	-.0035
#5	4.310	2.492	-.0043	4.175	-.0023

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.7213
Stddev	.2195
%RSD	2.8425

#1	7.9336
#2	7.7157
#3	7.9330
#4	7.5998
#5	7.4246

Method: ISO_U Sample Name: icv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/18/99 13:33 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.783	4.463	42.16	4.516	4955.
Stddev	.637	.197	1.69	.387	50.
%RSD	16.85	4.422	4.005	8.564	1.006
#1	4.398	4.201	43.72	4.726	5014.
#2	3.685	4.343	43.17	4.014	4911.
#3	3.998	4.465	41.85	4.465	4948.
#4	2.737	4.674	42.66	4.340	4905.
#5	4.094	4.631	39.40	5.034	4999.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5315
Stddev	.0728
%RSD	.96687

#1	7.6173
#2	7.5985
#3	7.5035
#4	7.4888
#5	7.4493

HNF-1663 REV. 0

Method: ISO_U Sample Name: icv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/18/99 13:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.783	4.463	.1453	4.516	21.19
Stddev	.637	.197	.0065	.387	.21
%RSD	16.85	4.422	4.499	8.564	1.007
#1	4.398	4.201	.1514	4.726	21.44
#2	3.685	4.343	.1493	4.014	21.00
#3	3.998	4.465	.1441	4.465	21.16
#4	2.737	4.674	.1473	4.340	20.97
#5	4.094	4.631	.1347	5.034	21.38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5315
Stddev	.0728
%RSD	.96687

#1	7.6173
#2	7.5985
#3	7.5035
#4	7.4888
#5	7.4493

Method: ISO_U

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/18/99 13:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.589	3.423	3.784	3.662	4.049
Stddev	.819	.292	.472	.434	.498
%RSD	22.81	8.530	12.48	11.87	12.30
#1	2.420	3.227	4.436	3.159	4.302
#2	3.051	3.496	3.178	3.560	3.878
#3	4.251	3.285	3.800	3.800	4.766
#4	3.973	3.903	3.973	4.322	3.834
#5	4.252	3.206	3.533	3.467	3.467

Int. Std.

Ir[193]

Line	193/pulse
Units	Cts/S
Avg	7.5767
Stddev	.2759
%RSD	3.6415

#1	7.4383
#2	7.8656
#3	7.7634
#4	7.1733
#5	7.6427

Method: ISO_U Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/18/99 13:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.589	3.423	-.0033	3.662	-.0025
Stddev	.819	.292	.0018	.434	.0021
%RSD	22.81	8.530	56.24	11.87	84.93

#1	2.420	3.227	-.0007	3.159	-.0014
#2	3.051	3.496	-.0056	3.560	-.0032
#3	4.251	3.285	-.0032	3.800	.0006
#4	3.973	3.903	-.0025	4.322	-.0034
#5	4.252	3.206	-.0042	3.467	-.0050

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5767
Stddev	.2759
%RSD	3.6415

#1	7.4383
#2	7.8656
#3	7.7634
#4	7.1733
#5	7.6427

Method: ISO_U

Sample Name: s98t002968

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:41 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.967	3.752	4.121	3.800	8.644
Stddev	.431	.695	.482	.462	1.097
%RSD	10.86	18.53	11.69	12.17	12.69
#1	3.508	3.444	4.720	3.763	8.675
#2	4.143	4.844	4.207	3.888	7.904
#3	4.172	3.642	3.377	3.046	10.40
#4	3.529	2.964	4.094	4.023	7.552
#5	4.484	3.863	4.208	4.277	8.693

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5132
Stddev	.3434
%RSD	4.5708

#1	7.8390
#2	7.8440
#3	7.5512
#4	7.0841
#5	7.2476

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t002968

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.967	3.752	-.0019	3.800	.0172
Stddev	.431	.695	.0019	.462	.0047
%RSD	10.86	18.53	95.88	12.17	27.36
#1	3.508	3.444	.0004	3.763	.0173
#2	4.143	4.844	-.0016	3.888	.0140
#3	4.172	3.642	-.0048	3.046	.0246
#4	3.529	2.964	-.0021	4.023	.0125
#5	4.484	3.863	-.0016	4.277	.0174

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5132
Stddev	.3434
%RSD	4.5708

#1	7.8390
#2	7.8440
#3	7.5512
#4	7.0841
#5	7.2476

HNF-1663 REV. 0

Method: ISO_U   Sample Name: s98t002968_d   Operator: bjg
Comment: 40501 df
Run Time: 01/18/99 13:43 Type: Unk   Mode: IR   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.695	3.448	3.577	3.340	8.083
Stddev	.586	.703	.921	.253	.538
%RSD	15.87	20.40	25.75	7.565	6.651
#1	4.249	3.203	4.510	3.530	7.452
#2	2.792	3.989	4.654	2.925	8.376
#3	4.055	2.991	2.859	3.523	8.841
#4	3.441	4.367	2.845	3.441	7.808
#5	3.937	2.691	3.019	3.281	7.940

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5736
Stddev	.0583
%RSD	.76932

#1	7.6493
#2	7.5212
#3	7.5214
#4	7.5568
#5	7.6193

Method: ISO_U Sample Name: s98t002968_d Operator: bjg
Comment: 40501 df
Run Time: 01/18/99 13:43 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.695	3.448	-.0041	3.340	.0148
Stddev	.586	.703	.0036	.253	.0023
%RSD	15.87	20.40	88.02	7.565	15.59
#1	4.249	3.203	-.0004	3.530	.0121
#2	2.792	3.989	.0001	2.925	.0160
#3	4.055	2.991	-.0068	3.523	.0180
#4	3.441	4.367	-.0069	3.441	.0136
#5	3.937	2.691	-.0062	3.281	.0141

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5736
Stddev	.0583
%RSD	.76932

#1	7.6493
#2	7.5212
#3	7.5214
#4	7.5568
#5	7.6193

Method: ISO_U

Sample Name: s98t002968_a

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:45 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.579	3.672	38.52	4.453	4553.
Stddev	.966	.470	1.78	.355	62.
%RSD	26.99	12.79	4.611	7.997	1.366
#1	2.688	3.868	40.32	4.851	4498.
#2	2.891	3.497	37.86	4.169	4479.
#3	3.064	3.530	36.43	3.995	4625.
#4	4.558	3.105	40.43	4.624	4573.
#5	4.694	4.358	37.55	4.625	4591.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.5191
Stddev	.0791
%RSD	1.0516

#1	7.6266
#2	7.4356
#3	7.5071
#4	7.5690
#5	7.4570

Method: ISO_U Sample Name: s98t002968_a Operator: bjg
Comment: 40501 df
Run Time: 01/18/99 13:45 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.579	3.672	.1312	4.453	19.47
Stddev	.966	.470	.0069	.356	.27
%RSD	26.99	12.79	5.241	7.997	1.367
#1	2.688	3.868	.1382	4.851	19.23
#2	2.891	3.497	.1287	4.169	19.15
#3	3.064	3.530	.1232	3.996	19.77
#4	4.558	3.105	.1386	4.624	19.55
#5	4.694	4.358	.1275	4.626	19.63

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	7.5191
Stddev	.0791
%RSD	1.0516

#1	7.6266
#2	7.4356
#3	7.5071
#4	7.5690
#5	7.4570

Method: ISO_U Sample Name: s98t002969

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.330	3.848	3.604	3.472	6.348
Stddev	.403	.289	.262	.442	.491
%RSD	12.11	7.515	7.259	12.73	7.727
#1	3.680	3.997	3.299	3.489	7.106
#2	3.772	3.896	3.587	2.905	6.431
#3	3.315	3.434	3.434	3.730	5.802
#4	2.819	3.716	3.973	4.037	6.343
#5	3.063	4.196	3.729	3.197	6.060

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9447
Stddev	.3485
%RSD	4.3869

#1	7.8808
#2	8.0855
#3	8.4459
#4	7.8036
#5	7.5078

Method: ISO_U   Sample Name: s98t002969   Operator: bjg
Comment: 40501 df
Run Time: 01/18/99 13:53 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.330	3.848	-.0039	3.472	.0073
Stddev	.403	.289	.0010	.442	.0021
%RSD	12.11	7.515	25.67	12.73	28.64
#1	3.680	3.997	-.0051	3.489	.0106
#2	3.772	3.896	-.0040	2.906	.0077
#3	3.315	3.434	-.0046	3.730	.0050
#4	2.819	3.716	-.0025	4.037	.0073
#5	3.063	4.196	-.0035	3.197	.0061

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9447
Stddev	.3485
%RSD	4.3869

#1	7.8808
#2	8.0855
#3	8.4459
#4	7.8036
#5	7.5078

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t002969_d

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.852	3.461	3.738	3.288	6.447
Stddev	.421	.223	.687	.475	.990
%RSD	10.93	6.443	18.38	14.44	15.35
#1	4.174	3.551	3.115	3.427	5.545
#2	3.906	3.162	3.968	3.224	5.828
#3	4.242	3.619	2.932	3.993	7.674
#4	3.745	3.293	4.132	3.099	7.360
#5	3.190	3.681	4.540	2.699	5.828

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9996
Stddev	.1525
%RSD	1.9058

#1	8.0252
#2	8.0640
#3	8.0143
#4	7.7441
#5	8.1501

Method: ISO_U   Sample Name: s98t002969_d   Operator: bjg
Comment: 40501 df
Run Time: 01/18/99 13:55 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.852	3.461	-.0034	3.288	.0078
Stddev	.421	.223	.0027	.475	.0042
%RSD	10.93	6.443	77.55	14.44	54.64
#1	4.174	3.551	-.0058	3.427	.0039
#2	3.906	3.162	-.0025	3.224	.0051
#3	4.242	3.619	-.0065	3.993	.0130
#4	3.745	3.293	-.0019	3.099	.0117
#5	3.190	3.681	-.0003	2.699	.0051

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9996
Stddev	.1525
%RSD	1.9058

#1	8.0252
#2	8.0640
#3	8.0143
#4	7.7441
#5	8.1501

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003010

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99

13:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.981	3.192	3.318	3.549	6.596
Stddev	.819	.548	.312	.849	.171
%RSD	20.58	17.16	9.416	23.93	2.597
#1	4.225	3.199	3.259	3.923	6.578
#2	5.226	4.106	3.048	3.360	6.408
#3	3.928	2.762	3.130	3.867	6.506
#4	3.121	3.121	3.308	2.185	6.866
#5	3.405	2.774	3.846	4.414	6.621

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.0817
Stddev	.1376
%RSD	1.7023

#1	8.2850
#2	8.0368
#3	8.1465
#4	8.0104
#5	7.9298

Method: ISO_U Sample Name: s98t003010

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.981	3.192	-.0051	3.549	.0084
Stddev	.819	.548	.0012	.849	.0007
%RSD	20.58	17.16	23.94	23.93	8.738
#1	4.225	3.199	-.0053	3.923	.0083
#2	5.226	4.106	-.0061	3.360	.0076
#3	3.928	2.762	-.0058	3.867	.0080
#4	3.121	3.121	-.0051	2.185	.0095
#5	3.405	2.774	-.0030	4.414	.0085

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.0817
Stddev	.1376
%RSD	1.7023

#1	8.2850
#2	8.0368
#3	8.1465
#4	8.0104
#5	7.9298

Method: ISO_U

Sample Name: s98t003010_d

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99

13:59 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.442	3.865	3.538	3.363	7.097
Stddev	.486	.603	.594	.432	.850
%RSD	14.11	15.61	16.79	12.86	11.98
#1	3.553	3.794	4.035	3.433	6.264
#2	4.111	4.423	4.236	3.426	7.288
#3	3.192	4.068	3.505	2.691	8.449
#4	3.551	4.174	2.991	3.364	6.542
#5	2.802	2.863	2.924	3.899	6.945

Int. Std.

Line	193/pulse
Units	Cts/S
Avg	8.1101
Stddev	.1369
%RSD	1.6878

#1	8.3020
#2	8.0271
#3	7.9893
#4	8.0251
#5	8.2071

Method: ISO_U

Sample Name: s98t003010_d

Operator: bjg

Comment: 40501 df

Run Time: 01/18/99 13:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.442	3.865	-.0042	3.363	.0105
Stddev	.486	.603	.0023	.432	.0036
%RSD	14.11	15.61	54.72	12.86	34.53
#1	3.553	3.794	-.0023	3.433	.0070
#2	4.111	4.423	-.0015	3.426	.0113
#3	3.192	4.068	-.0043	2.691	.0163
#4	3.551	4.174	-.0063	3.364	.0082
#5	2.802	2.863	-.0066	3.899	.0099

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.1101
Stddev	.1369
%RSD	1.6878

#1	8.3020
#2	8.0271
#3	7.9893
#4	8.0251
#5	8.2071

Method: ISO_U Sample Name: ccv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/18/99 14:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.677	4.295	43.15	4.634	5172.
Stddev	.695	.671	2.54	.552	228.
%RSD	18.91	15.62	5.891	11.91	4.398
#1	4.492	4.152	46.08	5.513	5562.
#2	3.494	5.241	45.24	4.368	5151.
#3	4.089	3.586	42.53	4.593	5129.
#4	2.645	3.814	42.14	4.675	5022.
#5	3.664	4.685	39.76	4.024	4998.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9521
Stddev	.3679
%RSD	4.6269

#1	7.3460
#2	8.0136
#3	7.9472
#4	8.1285
#5	8.3253

Method: ISO_U Sample Name: ccv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/18/99 14:03 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.677	4.295	.1492	4.634	22.12
Stddev	.695	.671	.0098	.552	.97
%RSD	18.91	15.62	6.598	11.91	4.402
#1	4.492	4.152	.1605	5.513	23.78
#2	3.494	5.241	.1573	4.368	22.03
#3	4.089	3.586	.1468	4.593	21.93
#4	2.645	3.814	.1453	4.675	21.47
#5	3.664	4.685	.1361	4.024	21.37

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.9521
Stddev	.3679
%RSD	4.6269

#1	7.3460
#2	8.0136
#3	7.9472
#4	8.1285
#5	8.3253

Method: ISO_U Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/18/99 14:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.157	3.473	3.512	3.288	4.243
Stddev	.598	.473	.756	.317	.550
%RSD	18.93	13.63	21.52	9.655	12.96

#1	2.419	3.776	3.009	2.891	4.012
#2	2.852	3.980	3.208	3.149	4.515
#3	3.474	3.174	4.552	3.234	4.132
#4	3.057	2.812	2.751	3.424	3.546
#5	3.982	3.620	4.042	3.740	5.007

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.3410
Stddev	.1149
%RSD	1.3780

#1	8.4741
#2	8.4164
#3	8.3487
#4	8.1781
#5	8.2878

Method: ISO_U Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/18/99 14:08 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	3.157	3.473	-.0043	3.288	-.0017
Stddev	.598	.473	.0029	.317	.0024
%RSD	18.93	13.63	68.03	9.655	139.9
#1	2.419	3.776	-.0063	2.891	-.0027
#2	2.852	3.980	-.0055	3.149	-.0005
#3	3.474	3.174	-.0003	3.234	-.0022
#4	3.057	2.812	-.0073	3.424	-.0047
#5	3.982	3.620	-.0023	3.740	.0016

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.3410
Stddev	.1149
%RSD	1.3780

#1	8.4741
#2	8.4164
#3	8.3487
#4	8.1781
#5	8.2878

LABCORE Completed Worklist Report for Worklist# 27504

Analyst: bjg

Instrument: ICPMS1

Book#: 42B57, 990119-3

Method: LA-506-101

Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		QMSPR-QC TC99	QC	0.0197	1.61e-02	81.726 % Recovery	
1 ICV	0		QMSPR-QC SN117	QC	0.0077	6.77e-03	87.922 % Recovery	
1 ICV	0		QMSPR-QC SN126	QC	n/a	<1.50e-4	% Recovery	
1 ICV	0		QMSPR-QC CS133	QC	0.0200	2.14e-02	107.000 % Recovery	
1 ICV	0		QMSPR-QC CS135	QC	n/a	3.33e-05	% Recovery	
1 ICV	0		QMSPR-QC CS137	QC	n/a	5.78e-05	% Recovery	
2 ICB	0		QMSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2 ICB	0		QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2 ICB	0		QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2 ICB	0		QMSPR-QC CS133	QC	1	1.06e-05	1.060e-005 ug/mL	
2 ICB	0		QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2 ICB	0		QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 TC99-D1 LIQUID		<u>N/A</u>	4.755e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 SN117-D1 LIQUID		<u>N/A</u> <	6.075e+00	6.075 ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 SN126-D1 LIQUID		<u>N/A</u> <	6.075e+00	6.075 ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 CS133-D1 LIQUID		<u>N/A</u>	7.310e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 CS135-D1 LIQUID		<u>N/A</u>	2.005e+00	8.10e-002 ug/mL	
3 SAMPLE	S98T003011	0 D	QMSPR-D1 CS137-D1 LIQUID		<u>N/A</u>	1.701e+00	8.10e-002 ug/mL	
4 DUP	S98T003011	0 D	QMSPR-D1 TC99-D1 LIQUID		4.75e+00	4.91e+00	3.313 RPD	
4 DUP	S98T003011	0 D	QMSPR-D1 SN117-D1 LIQUID		<6.08e0	<6.08e0	RPD	
4 DUP	S98T003011	0 D	QMSPR-D1 SN126-D1 LIQUID		<6.08e0	<6.08e0	RPD	
4 DUP	S98T003011	0 D	QMSPR-D1 CS133-D1 LIQUID		7.31e+00	7.57e+00	3.495 RPD	
4 DUP	S98T003011	0 D	QMSPR-D1 CS135-D1 LIQUID		2.00e+00	2.19e+00	9.069 RPD	
4 DUP	S98T003011	0 D	QMSPR-D1 CS137-D1 LIQUID		1.70e+00	1.95e+00	13.699 RPD	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 TC99-D1 LIQUID		<u>N/A</u>	5.196e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 SN117-D1 LIQUID		<u>N/A</u> <	6.075e+00	6.075 ug/mL	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 SN126-D1 LIQUID		<u>N/A</u> <	6.075e+00	6.075 ug/mL	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 CS133-D1 LIQUID		<u>N/A</u>	7.930e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 CS135-D1 LIQUID		<u>N/A</u>	2.398e+00	8.10e-002 ug/mL	
5 SAMPLE	S98T003012	0 D	QMSPR-D1 CS137-D1 LIQUID		<u>N/A</u>	1.993e+00	8.10e-002 ug/mL	
6 DUP	S98T003012	0 D	QMSPR-D1 TC99-D1 LIQUID		5.20e+00	5.22e+00	0.384 RPD	
6 DUP	S98T003012	0 D	QMSPR-D1 SN117-D1 LIQUID		<6.08e0	<6.08e0	RPD	
6 DUP	S98T003012	0 D	QMSPR-D1 SN126-D1 LIQUID		<6.08e0	<6.08e0	RPD	
6 DUP	S98T003012	0 D	QMSPR-D1 CS133-D1 LIQUID		7.93e+00	7.67e+00	3.333 RPD	
6 DUP	S98T003012	0 D	QMSPR-D1 CS135-D1 LIQUID		2.40e+00	2.27e+00	5.567 RPD	
6 DUP	S98T003012	0 D	QMSPR-D1 CS137-D1 LIQUID		1.99e+00	1.90e+00	4.627 RPD	
7 SPK-POST	S98T003012	0 D	QMSPR-D1 TC99-D1 LIQUID		0.0219	1.80e-02	82.192 % Recovery	
7 SPK-POST	S98T003012	0 D	QMSPR-D1 SN117-D1 LIQUID		n/a	4.65e-03	% Recovery	
7 SPK-POST	S98T003012	0 D	QMSPR-D1 SN126-D1 LIQUID		n/a	1.10e-06	% Recovery	
7 SPK-POST	S98T003012	0 D	QMSPR-D1 CS133-D1 LIQUID		0.0201	2.18e-02	108.458 % Recovery	
7 SPK-POST	S98T003012	0 D	QMSPR-D1 CS135-D1 LIQUID		n/a	-9.00e-7	% Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual

LABCORE Completed Worklist Report for Worklist# 27504

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 SPK-POST	S98T003012	0 D	QMSPR-D1 CS137-D1	LIQUID	n/a	3.00e-07		% Recovery
8 SAMPLE	S98T003013	0 D	QMSPR-D1 TC99-D1	LIQUID	N/A	5.419e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003013	0 D	QMSPR-D1 SN117-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
8 SAMPLE	S98T003013	0 D	QMSPR-D1 SN126-D1	LIQUID	N/A	6.075e+00	6.075	ug/mL
8 SAMPLE	S98T003013	0 D	QMSPR-D1 CS133-D1	LIQUID	N/A	7.562e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003013	0 D	QMSPR-D1 CS135-D1	LIQUID	N/A	2.292e+00	8.10e-002	ug/mL
8 SAMPLE	S98T003013	0 D	QMSPR-D1 CS137-D1	LIQUID	N/A	1.912e+00	8.10e-002	ug/mL
9 DUP	S98T003013	0 D	QMSPR-D1 TC99-D1	LIQUID	5.42e+00	5.33e+00	1.674	RPD
9 DUP	S98T003013	0 D	QMSPR-D1 SN117-D1	LIQUID	<6.08e0	<6.08e0		RPD
9 DUP	S98T003013	0 D	QMSPR-D1 SN126-D1	LIQUID	<6.08e0	<6.08e0		RPD
9 DUP	S98T003013	0 D	QMSPR-D1 CS133-D1	LIQUID	7.56e+00	7.50e+00	0.797	RPD
9 DUP	S98T003013	0 D	QMSPR-D1 CS135-D1	LIQUID	2.29e+00	2.24e+00	2.208	RPD
9 DUP	S98T003013	0 D	QMSPR-D1 CS137-D1	LIQUID	1.91e+00	1.95e+00	2.073	RPD
10 CCV	0		QMSPR-QC TC99	QC	0.0197	1.62e-02	82.234	% Recovery
10 CCV	0		QMSPR-QC SN117	QC	0.0077	6.72e-03	87.273	% Recovery
10 CCV	0		QMSPR-QC SN126	QC	n/a	<1.50e-4		% Recovery
10 CCV	0		QMSPR-QC CS133	QC	0.0200	2.14e-02	107.000	% Recovery
10 CCV	0		QMSPR-QC CS135	QC	n/a	3.46e-05		% Recovery
10 CCV	0		QMSPR-QC CS137	QC	n/a	5.99e-05		% Recovery
11 CCB	0		QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
11 CCB	0		QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
11 CCB	0		QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
11 CCB	0		QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27504

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. Pang 1/21/99
Reviewer Signature _____ Date _____

12/10/98 06:28
A-0004-1

FILED PRO120A.TPT

Page: 1

LABCORE Data Entry Template for Worklist# 27504

Analyst: B. GORR Instrument: ICPMS1 _____ Book# 44857, 990119-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC	QC		
2 ICB			@MSPR-QC	QC		
3 SAMPLE	S98T003011 0 D		@MSPR-D1 LIQUID		98000510	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
4 DUP	S98T003011 0 D		@MSPR-D1 LIQUID			
5 SAMPLE	S98T003012 0 D		@MSPR-D1 LIQUID		98000510	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
6 DUP	S98T003012 0 D		@MSPR-D1 LIQUID			
7 SPK-Post	S98T003012 0 D		@MSPR-D1 LIQUID			
8 SAMPLE	S98T003013 0 D		@MSPR-D1 LIQUID		98000510	BY-105 (2)
Analytes Requested: CS133-D1, CS135-D1, CS137-D1, I129-D1, SN117-D1, SN126-D1, TC99-D1						
9 DUP	S98T003013 0 D		@MSPR-D1 LIQUID			
10 CCV			@MSPR-QC	QC		
11 CCB			@MSPR-QC	QC		

Final page for worklist # 27504

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

1-20-99
S98T003011, 3012, 3013
1-10-025-10 99501

POST SPIKE = .1 ml of MSPE. spk. 44857
 in 10ml
 Note: Sn was not spike

SCANS = 983011, PR3012, PR3013

Upgraded & Validated by
Sal M. Pang

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

1/21/99

2533

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/20/99

10:37 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	115400.
Stddev	647.
%RSD	.5605

#1	115400.
#2	115100.
#3	115400.
#4	114900.
#5	114600.
#6	114900.
#7	115100.
#8	116400.
#9	115700.
#10	116600.

Bjg
1-20-99

BY-105
588700 3011-13
1-10-025-10

FIRST PR 0120A RT
W.L.H. 27504

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2534 TO 2551

Method: TC_SN_CS Sample Name: Blank

Operator:

Comment:

Run Time: 01/20/99 10:44 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.833	1.856	9.504	1.988	.9943
Stddev	.251	.475	.570	.351	.1152
%RSD	13.72	25.58	5.993	17.68	11.59

#1	1.841	1.800	9.778	2.168	1.146
#2	2.039	1.872	9.361	1.706	1.082
#3	1.979	1.113	8.988	1.567	.9482
#4	1.404	2.105	9.040	2.064	.8669
#5	1.902	2.389	10.35	2.433	.9291

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	10.19	11.79	19.03	117.2
Stddev	.67	1.07	.81	3.8
%RSD	6.574	9.071	4.260	3.267

#1	9.532	10.43	19.56	111.9
#2	9.693	11.19	17.97	118.7
#3	11.01	12.20	19.95	117.0
#4	9.907	11.85	19.24	116.2
#5	10.80	13.27	18.45	122.4

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.956
Stddev	.373
%RSD	3.1209

#1	12.221
#2	12.019
#3	12.128
#4	12.112
#5	11.301

Method: TC_SN_CS Sample Name: PRIV Std

Operator:

Comment:

Run Time: 01/20/99 10:49 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]
Line	99/pulse	100/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	12630.	12750.	6888.	6888.
Stddev	270.	250.	122.	122.
%RSD	2.139	1.961	1.765	1.765

#1	12890.	12870.	6804.	6804.
#2	12630.	12610.	6777.	6777.
#3	12400.	12460.	6865.	6865.
#4	12320.	12690.	6912.	6912.
#5	12900.	13110.	7084.	7084.

Elem	Te[125]	Cs[133]	Cs[135]	Cs[137]
Line	125/pulse	133/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1551.	83490.	5911.	10200.
Stddev	25.	785.	52.	172.
%RSD	1.632	.9402	.8880	1.684

#1	1560.	83540.	5856.	10160.
#2	1542.	84120.	5889.	10180.
#3	1518.	82240.	5900.	10000.
#4	1551.	83360.	5912.	10170.
#5	1587.	84190.	5997.	10480.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	61490.
Stddev	1543.
%RSD	2.510

#1	60470.
#2	61100.
#3	60000.
#4	61970.
#5	63920.

Sample Name: PRIV Std Run Time: 01/20/99 10:49

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.781
Stddev	.217
%RSD	2.0152

#1	10.795
#2	10.855
#3	11.004
#4	10.830
#5	10.419

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/20/99 10:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0042	.0004	6.766	.0005	-.0001
Stddev	.0006	.0005	.075	.0002	.0010
%RSD	14.74	123.4	1.102	38.99	771.3

#1	.0037	.0011	6.696	.0008	.0006
#2	.0053	.0006	6.853	.0005	.0004
#3	.0040	-.0001	6.827	.0003	-.0002
#4	.0042	.0003	6.765	.0006	-.0019
#5	.0040	.0001	6.688	.0004	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0180	-.0086	-.0166	-.0120
Stddev	.0014	.0006	.0003	.0001
%RSD	7.500	7.126	1.546	.7493
#1	.0170	-.0090	-.0169	-.0119
#2	.0204	-.0083	-.0165	-.0120
#3	.0175	-.0089	-.0168	-.0119
#4	.0179	-.0093	-.0162	-.0119
#5	.0174	-.0077	-.0166	-.0121

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.171
Stddev	.210
%RSD	1.8830

#1	11.525
#2	11.030
#3	11.011
#4	11.098
#5	11.191

HNF-1663 REV.0

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/20/99 11:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.05	.0064	1.347	.0001	-.0003
Stddev	.34	.0002	.046	.0002	.0010
%RSD	2.093	3.244	3.424	320.3	306.9

#1	15.70	.0063	1.266	.0001	.0004
#2	15.85	.0063	1.359	-.0001	-.0006
#3	16.07	.0065	1.371	.0005	.0005
#4	16.07	.0063	1.363	.0000	.0001
#5	16.58	.0068	1.377	-.0001	-.0021

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.38	.0333	.0578	.0400
Stddev	.36	.0020	.0027	.0012
%RSD	1.678	5.911	4.662	3.088

#1	21.12	.0343	.0541	.0395
#2	21.35	.0324	.0560	.0391
#3	21.49	.0361	.0595	.0390
#4	21.03	.0309	.0585	.0418
#5	21.94	.0330	.0608	.0408

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.386
Stddev	.192
%RSD	1.6874

#1	11.598
#2	11.478
#3	11.435
#4	11.329
#5	11.089

Method: TC_SN_CS Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/20/99 11:08 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	-.0004	.0010	.0003	-.0013
Stddev	.0002	.0002	.0006	.0003	.0008
%RSD	41.99	45.92	58.67	104.6	63.72

#1	.0005	-.0002	.0020	-.0001	-.0021
#2	.0004	-.0005	.0006	.0004	.0000
#3	.0007	-.0005	.0009	.0006	-.0014
#4	.0009	-.0002	.0005	.0004	-.0018
#5	.0003	-.0004	.0011	.0000	-.0010

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0106	.0002	.0002	.0000
Stddev	.0016	.0005	.0012	.000
%RSD	14.72	224.2	519.6	1662.

#1	.0094	-.0005	-.0008	-.0001
#2	.0128	.0005	.0012	.0005
#3	.0089	-.0002	-.0001	-.0003
#4	.0114	.0007	-.0009	.0000
#5	.0104	.0006	.0017	-.0003

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.783
Stddev	.244
%RSD	2.0724

#1	12.210
#2	11.635
#3	11.720
#4	11.733
#5	11.617

Method: TC_SN_CS Sample Name: s98t003011

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1174	.1472	.0437	.0010	-.0005
Stddev	.0029	.0016	.0015	.0003	.0006
%RSD	2.456	1.109	3.370	29.89	111.5

#1	.1153	.1492	.0415	.0008	-.0005
#2	.1200	.1469	.0442	.0015	-.0004
#3	.1138	.1447	.0446	.0010	-.0011
#4	.1173	.1475	.0428	.0010	.0004
#5	.1205	.1475	.0451	.0008	-.0010

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1805	.0495	.0420	-.0121
Stddev	.0029	.0017	.0015	.0001
%RSD	1.597	3.409	3.458	.6679

#1	.1807	.0474	.0419	-.0122
#2	.1795	.0519	.0407	-.0121
#3	.1763	.0502	.0405	-.0121
#4	.1818	.0495	.0427	-.0121
#5	.1841	.0486	.0440	-.0120

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.713
Stddev	.196
%RSD	1.6774

#1	11.905
#2	11.632
#3	11.936
#4	11.594
#5	11.496

Method: TC_SN_CS Sample Name: s98t003011_d

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1212	.1505	.0459	.0011	-.0004
Stddev	.0020	.0043	.0020	.0005	.0008
%RSD	1.641	2.872	4.325	46.53	184.0

#1	.1234	.1529	.0445	.0015	-.0001
#2	.1194	.1529	.0437	.0012	-.0001
#3	.1197	.1439	.0464	.0005	-.0004
#4	.1202	.1485	.0458	.0006	-.0017
#5	.1233	.1544	.0489	.0015	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1868	.0540	.0481	-.0092
Stddev	.0028	.0018	.0018	.0002
%RSD	1.477	3.351	3.641	1.904

#1	.1854	.0549	.0457	-.0093
#2	.1845	.0512	.0472	-.0094
#3	.1859	.0541	.0482	-.0091
#4	.1865	.0536	.0490	-.0090
#5	.1915	.0560	.0503	-.0090

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	11.160
Stddev	.180
%RSD	1.6109

#1	11.259
#2	11.394
#3	11.081
#4	11.143
#5	10.919

Method: TC_SN_CS Sample Name: s98t003012

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1283	.1558	.0480	.0018	-.0009
Stddev	.0029	.0041	.0024	.0004	.0008
%RSD	2.289	2.629	4.984	24.78	97.16

#1	.1309	.1514	.0442	.0014	.0001
#2	.1253	.1593	.0470	.0018	-.0008
#3	.1315	.1573	.0499	.0014	-.0021
#4	.1286	.1597	.0492	.0024	-.0012
#5	.1254	.1515	.0495	.0019	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1958	.0592	.0492	-.0122
Stddev	.0031	.0015	.0024	.0001
%RSD	1.571	2.506	4.882	.8060

#1	.1939	.0574	.0459	-.0122
#2	.1984	.0604	.0483	-.0122
#3	.1936	.0583	.0487	-.0121
#4	.1999	.0610	.0510	-.0120
#5	.1935	.0590	.0520	-.0123

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.130
Stddev	.176
%RSD	1.5809

#1	11.385
#2	11.128
#3	11.017
#4	10.925
#5	11.195

Method: TC_SN_CS Sample Name: s98t003012_d

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:25 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1290	.1576	.0486	.0012	-.0003
Stddev	.0032	.0032	.0031	.0004	.0009
%RSD	2.485	2.006	6.353	36.22	348.4

#1	.1320	.1587	.0445	.0008	-.0010
#2	.1257	.1600	.0510	.0007	.0006
#3	.1278	.1600	.0463	.0016	-.0012
#4	.1268	.1524	.0500	.0014	.0008
#5	.1329	.1570	.0514	.0016	-.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1894	.0560	.0469	-.0127
Stddev	.0036	.0020	.0023	.0001
%RSD	1.881	3.569	4.883	.6239

#1	.1852	.0530	.0432	-.0128
#2	.1920	.0569	.0470	-.0126
#3	.1865	.0569	.0470	-.0126
#4	.1937	.0549	.0480	-.0127
#5	.1894	.0581	.0494	-.0126

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	11.446
Stddev	.209
%RSD	1.8260

#1	11.723
#2	11.453
#3	11.451
#4	11.471
#5	11.134

Method: TC_SN_CS Sample Name: s98t003012_a

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.11	.1530	4.648	.0011	-.0005
Stddev	.26	.0029	.063	.0004	.0012
%RSD	1.463	1.923	1.350	33.06	232.1

#1	18.48	.1546	4.619	.0010	.0016
#2	17.88	.1485	4.573	.0015	-.0009
#3	18.12	.1517	4.624	.0014	-.0009
#4	17.83	.1561	4.695	.0006	-.0009
#5	18.21	.1541	4.729	.0009	-.0017

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.96	.0583	.0495	-.0115
Stddev	.23	.0018	.0023	.0001
%RSD	1.025	3.018	4.737	.9129

#1	22.25	.0606	.0507	-.0117
#2	21.89	.0558	.0524	-.0115
#3	21.89	.0584	.0503	-.0114
#4	21.67	.0578	.0475	-.0116
#5	22.12	.0590	.0467	-.0115

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.167
Stddev	.067
%RSD	.59749

#1	11.167
#2	11.130
#3	11.281
#4	11.144
#5	11.113

Method: TC_SN_CS Sample Name: s98t003013

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1338	.1601	.0480	.0016	-.0002
Stddev	.0018	.0057	.0013	.0007	.0006
%RSD	1.369	3.578	2.754	44.22	253.6

#1	.1325	.1548	.0483	.0017	-.0007
#2	.1342	.1606	.0456	.0015	.0002
#3	.1368	.1695	.0489	.0015	.0005
#4	.1322	.1562	.0485	.0006	-.0001
#5	.1334	.1594	.0486	.0025	-.0009

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1867	.0566	.0472	-.0125
Stddev	.0052	.0033	.0023	.0001
%RSD	2.785	5.858	4.863	.6585

#1	.1807	.0515	.0483	-.0125
#2	.1839	.0589	.0434	-.0126
#3	.1944	.0601	.0474	-.0126
#4	.1887	.0559	.0476	-.0125
#5	.1860	.0565	.0495	-.0124

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.255
Stddev	.209
%RSD	1.8573

#1	11.386
#2	11.196
#3	10.913
#4	11.411
#5	11.370

Method: TC_SN_CS Sample Name: s98t003013_d

Operator: bjg

Comment: 40501 df

Run Time: 01/20/99 11:40 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1316	.1590	.0488	.0015	-.0010
Stddev	.0021	.0031	.0015	.0004	.0006
%RSD	1.591	1.976	3.158	27.20	59.65

#1	.1292	.1594	.0510	.0015	-.0006
#2	.1298	.1561	.0495	.0017	-.0016
#3	.1328	.1566	.0472	.0019	-.0011
#4	.1319	.1589	.0476	.0009	-.0016
#5	.1342	.1640	.0490	.0012	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.1851	.0552	.0482	-.0127
Stddev	.0056	.0042	.0019	.0000
%RSD	3.018	7.550	3.872	.2974

#1	.1855	.0516	.0504	-.0127
#2	.1813	.0501	.0452	-.0127
#3	.1782	.0576	.0487	-.0126
#4	.1926	.0570	.0483	-.0127
#5	.1877	.0599	.0485	-.0127

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.335
Stddev	.317
%RSD	2.7997

#1	11.558
#2	11.733
#3	11.301
#4	10.938
#5	11.144

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/20/99 11:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0073	-.0003	6.721	.0005	.0004
Stddev	.0011	.0003	.099	.0004	.0010
%RSD	14.57	101.6	1.467	81.39	246.5

#1	.0069	-.0003	6.551	.0002	.0002
#2	.0092	-.0005	6.781	.0002	.0011
#3	.0070	-.0003	6.742	.0010	.0010
#4	.0070	.0002	6.797	.0009	-.0013
#5	.0065	-.0004	6.733	.0002	.0010

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0026	-.0091	-.0168	-.0123
Stddev	.0008	.0006	.0003	.0001
%RSD	30.75	6.058	1.537	1.020

#1	-.0022	-.0094	-.0170	-.0124
#2	-.0016	-.0082	-.0172	-.0122
#3	-.0022	-.0095	-.0167	-.0123
#4	-.0035	-.0095	-.0166	-.0122
#5	-.0033	-.0090	-.0168	-.0124

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.691
Stddev	.191
%RSD	1.7890

#1	10.999
#2	10.522
#3	10.642
#4	10.556
#5	10.737

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/20/99 11:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	16.16	.0069	1.393	.0000	.0003
Stddev	.29	.0009	.079	.000	.0017
%RSD	1.773	13.18	5.688	1139.	588.4

#1	16.22	.0072	1.431	-.0006	-.0005
#2	15.76	.0080	1.505	.0003	-.0018
#3	16.44	.0072	1.386	-.0001	.0019
#4	16.41	.0068	1.298	.0004	.0023
#5	15.99	.0055	1.347	-.0002	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.37	.0346	.0599	.0396
Stddev	.17	.0015	.0027	.0007
%RSD	.7900	4.458	4.514	1.646

#1	21.29	.6329	.0602	.0404
#2	21.41	.0351	.0591	.0400
#3	21.48	.0369	.0623	.0397
#4	21.54	.0346	.0557	.0388
#5	21.11	.0336	.0621	.0392

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.789
Stddev	.075
%RSD	.69781

#1	10.861
#2	10.815
#3	10.676
#4	10.753
#5	10.842

Method: TC_SN_CS Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/20/99 11:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0005	-.0003	.0000	.0002	.0001
Stddev	.0003	.0001	.001	.0004	.0005
%RSD	67.00	19.96	43950.	210.1	916.7

#1	.0009	-.0003	.0006	-.0001	-.0004
#2	.0003	-.0003	.0003	-.0002	-.0001
#3	.0008	-.0003	.0006	.0001	-.0002
#4	.0002	-.0002	-.0012	.0005	.0002
#5	.0003	-.0003	-.0003	.0006	.0009

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0026	-.0003	.0014	.0003
Stddev	.0006	.0005	.0011	.0003
%RSD	23.30	178.6	81.00	107.8

#1	-.0032	.0002	.0024	-.0001
#2	-.0018	.0004	-.0003	.0003
#3	-.0029	-.0008	.0023	.0007
#4	-.0022	-.0004	.0017	.0001
#5	-.0031	-.0007	.0008	.0005

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.033
Stddev	.127
%RSD	1.1496

#1	11.149
#2	10.905
#3	10.960
#4	11.188
#5	10.960

Method: TC_SN_CS Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/20/99 11:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	-.0003	.0001	.0000	-.0003
Stddev	.0003	.0002	.0010	.000	.0011
%RSD	281.6	90.08	1138.	632.0	412.2

#1	-.0001	-.0005	-.0011	.0003	.0009
#2	.0002	-.0004	.0012	-.0004	-.0013
#3	-.0001	-.0001	.0000	.0001	.0009
#4	.0007	-.0003	-.0007	-.0002	-.0010
#5	-.0002	.0000	.0010	.0001	-.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0038	-.0004	.0001	.0000
Stddev	.0005	.0013	.0010	.000
%RSD	13.20	313.7	1550.	1663.

#1	-.0030	.0009	-.0015	-.0004
#2	-.0037	-.0007	.0008	-.0002
#3	-.0043	-.0018	-.0001	.0000
#4	-.0041	.0009	.0011	.0003
#5	-.0037	-.0014	.0001	.0002

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	11.091
Stddev	.199
%RSD	1.7922

#1	11.281
#2	11.315
#3	10.979
#4	10.857
#5	11.022

LABCORE Completed Worklist Report for Worklist# 27505

Analyst: bjg BJG

Instrument: ICPMS1

Book#: 990104-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

Seq Type	Sample#	RA	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSU-QC U235	QC	0.000143	1.47e-04	102.797 % Recovery	
1	ICV	0	EMSU-QC U238	QC	0.020	2.04e-02	102.000 % Recovery	
2	ICB	0	EMSU-QC U233	QC	1	<5.19e-5		ug/mL
2	ICB	0	EMSU-QC U234	QC	1	<2.79e-5		ug/mL
2	ICB	0	EMSU-QC U235	QC	1	<2.79e-5		ug/mL
2	ICB	0	EMSU-QC U236	QC	1	<3.14e-5		ug/mL
2	ICB	0	EMSU-QC U238	QC	1	<1.90e-5		ug/mL
3	SAMPLE	S98T003011	0 D EMSU-D1 U233-D1	LIQUID	N/A	< 2.101e+00	2.101	ug/mL
3	SAMPLE	S98T003011	0 D EMSU-D1 U234-D1	LIQUID	N/A	< 1.128e+00	1.128	ug/mL
3	SAMPLE	S98T003011	0 D EMSU-D1 U235-D1	LIQUID	N/A	< 1.132e+00	1.132	ug/mL
3	SAMPLE	S98T003011	0 D EMSU-D1 U236-D1	LIQUID	N/A	< 1.271e+00	1.271	ug/mL
3	SAMPLE	S98T003011	0 D EMSU-D1 U238-D1	LIQUID	N/A	< 7.678e-01	0.768	ug/mL
4	DUP	S98T003011	0 D EMSU-D1 U233-D1	LIQUID	<2.10e0	<2.10e0		RPD
4	DUP	S98T003011	0 D EMSU-D1 U234-D1	LIQUID	<1.13e0	<1.13e0		RPD
4	DUP	S98T003011	0 D EMSU-D1 U235-D1	LIQUID	<1.13e0	<1.13e0		RPD
4	DUP	S98T003011	0 D EMSU-D1 U236-D1	LIQUID	<1.27e0	<1.27e0		RPD
4	DUP	S98T003011	0 D EMSU-D1 U238-D1	LIQUID	<7.68e-1	<7.68e-1		RPD
5	SAMPLE	S98T003012	0 D EMSU-D1 U233-D1	LIQUID	N/A	< 2.101e+00	2.101	ug/mL
5	SAMPLE	S98T003012	0 D EMSU-D1 U234-D1	LIQUID	N/A	< 1.128e+00	1.128	ug/mL
5	SAMPLE	S98T003012	0 D EMSU-D1 U235-D1	LIQUID	N/A	< 1.132e+00	1.132	ug/mL
5	SAMPLE	S98T003012	0 D EMSU-D1 U236-D1	LIQUID	N/A	< 1.271e+00	1.271	ug/mL
5	SAMPLE	S98T003012	0 D EMSU-D1 U238-D1	LIQUID	N/A	< 7.678e-01	0.768	ug/mL
6	DUP	S98T003012	0 D EMSU-D1 U233-D1	LIQUID	<2.10e0	<2.10e0		RPD
6	DUP	S98T003012	0 D EMSU-D1 U234-D1	LIQUID	<1.13e0	<1.13e0		RPD
6	DUP	S98T003012	0 D EMSU-D1 U235-D1	LIQUID	<1.13e0	<1.13e0		RPD
6	DUP	S98T003012	0 D EMSU-D1 U236-D1	LIQUID	<1.27e0	<1.27e0		RPD
6	DUP	S98T003012	0 D EMSU-D1 U238-D1	LIQUID	<7.68e-1	<7.68e-1		RPD
7	SPK-POST	S98T003012	0 D EMSU-D1 U233-D1	LIQUID	n/a	0.00e+00		% Recovery
7	SPK-POST	S98T003012	0 D EMSU-D1 U234-D1	LIQUID	n/a	0.00e+00		% Recovery
7	SPK-POST	S98T003012	0 D EMSU-D1 U235-D1	LIQUID	n/a	1.26e-04		% Recovery
7	SPK-POST	S98T003012	0 D EMSU-D1 U236-D1	LIQUID	n/a	-0.00e0		% Recovery
7	SPK-POST	S98T003012	0 D EMSU-D1 U238-D1	LIQUID	0.020	1.85e-02	92.500	% Recovery
8	SAMPLE	S98T003013	0 D EMSU-D1 U233-D1	LIQUID	N/A	< 2.101e+00	2.101	ug/mL
8	SAMPLE	S98T003013	0 D EMSU-D1 U234-D1	LIQUID	N/A	< 1.128e+00	1.128	ug/mL
8	SAMPLE	S98T003013	0 D EMSU-D1 U235-D1	LIQUID	N/A	< 1.132e+00	1.132	ug/mL
8	SAMPLE	S98T003013	0 D EMSU-D1 U236-D1	LIQUID	N/A	< 1.271e+00	1.271	ug/mL
8	SAMPLE	S98T003013	0 D EMSU-D1 U238-D1	LIQUID	N/A	< 7.678e-01	0.768	ug/mL
9	DUP	S98T003013	0 D EMSU-D1 U233-D1	LIQUID	<2.10e0	<2.10e0		RPD
9	DUP	S98T003013	0 D EMSU-D1 U234-D1	LIQUID	<1.13e0	<1.13e0		RPD
9	DUP	S98T003013	0 D EMSU-D1 U235-D1	LIQUID	<1.13e0	<1.13e0		RPD
9	DUP	S98T003013	0 D EMSU-D1 U236-D1	LIQUID	<1.27e0	<1.27e0		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27505

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
9	DUP	S98T003013	0	D	QMSU-D1 U238-D1	LIQUID	<7.68e-1	<7.68e-1		RPD
10	CCV		0		QMSU-QC U235	QC	0.000143	1.41e-04	98.601 % Recovery	
10	CCV		0		QMSU-QC U238	QC	0.020	2.01e-02	100.500 % Recovery	
11	CCB		0		QMSU-QC U233	QC	1	<5.19e-5		ug/mL
11	CCB		0		QMSU-QC U234	QC	1	<2.79e-5		ug/mL
11	CCB		0		QMSU-QC U235	QC	1	<2.79e-5		ug/mL
11	CCB		0		QMSU-QC U236	QC	1	<3.14e-5		ug/mL
11	CCB		0		QMSU-QC U238	QC	1	4.46e-03	0.004	ug/mL

Final page for worklist# 27505

_____
Analyst Signature   Date

_____
Analyst Signature   Date

Sam M. Pang 1/12/99
_____
Reviewer Signature   Date

LABCORE Data Entry Template for Worklist# 27505

Analyst: B. Gochet Instrument: ICPMS1 Book# 990104-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) MSU

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 SAMPLE	S98T003011 0 D		@MSU-D1	LIQUID	98000510 BY-105 (2)	
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
4 DUP	S98T003011 0 D		@MSU-D1	LIQUID		
5 SAMPLE	S98T003012 0 D		@MSU-D1	LIQUID	98000510 BY-105 (2)	
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
6 DUP	S98T003012 0 D		@MSU-D1	LIQUID		
7 SPK-POST	S98T003012 0 D		@MSU-D1	LIQUID		
8 SAMPLE	S98T003013 0 D		@MSU-D1	LIQUID	98000510 BY-105 (2)	
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
9 DUP	S98T003013 0 D		@MSU-D1	LIQUID		
10 CCV			@MSU-QC	QC		
11 CCB			@MSU-QC	QC		

Final page for worklist # 27505

Analyst Signature B. Gochet Date 1-11-99

Analyst Signature _____ Date _____

S98T003011
3011-D
3012-
3012-D
3012-A
3013
3013-D
Data Entry Comments:

DE
40501

POST spike = .02 ml of 10ppm U

Uploaded + Validated by:

Sam M. Pang

1/12/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune Operator: bjg
Comment:
Run Time: 01/11/99 10:37 Type: Unk Mode: INT Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	102600.
Stddev	1107.
%RSD	1.079

#1	102100.
#2	104000.
#3	102700.
#4	100100.
#5	102300.
#6	103400.
#7	102200.
#8	102100.
#9	103800.
#10	103200.

[Signature]
BY-105

ISO-U
FILE # 00111A-125
5887003011-13
.1-10-1025-10
Wt # 27505

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2555 TO 2583.

Method: ISO_U

Sample Name: BLANK

Operator:

Comment:

Run Time: 01/11/99 10:39 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U _[235]	U _[238]
Line	235/pulse	238/pulse
Units	Cts/S	Cts/S
Avg	5.834	5.543
Stddev	.580	.496
%RSD	9.947	8.948

#1	6.742	5.705
#2	5.922	5.076
#3	5.849	5.238
#4	5.288	6.328
#5	5.369	5.369

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	5.7927
Stddev	.0691
%RSD	1.1922

#1	5.7847
#2	5.9099
#3	5.7270
#4	5.7677
#5	5.7743

Method: ISO_U

Sample Name: U High Std

Operator:

Comment:

Run Time: 01/11/99

10:41 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[235]	U_[238]
Line	235/pulse	238/pulse
Units	Cts/S	Cts/S
Avg	196.0	24690.
Stddev	8.5	453.
%RSD	4.337	1.834

#1	209.5	24480.
#2	190.7	24500.
#3	191.3	24350.
#4	189.2	24660.
#5	199.3	25480.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	5.7255
Stddev	.0361
%RSD	.63056

#1	5.6904
#2	5.6936
#3	5.7790
#4	5.7335
#5	5.7312

Method: ISO_U Sample Name: iso-u std

Operator: bjg

Comment: 100 ppb

Run Time: 01/11/99 10:44 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.520	7.189	206.3	9.277	25570.
Stddev	.848	.338	13.2	.961	1155.
%RSD	15.36	4.696	6.394	10.36	4.519
#1	4.724	7.294	189.3	8.372	23830.
#2	6.063	7.240	226.2	10.86	26420.
#3	4.646	7.486	203.2	9.035	24940.
#4	6.608	6.608	206.2	9.391	25150.
#5	5.556	7.319	206.4	8.730	25500.

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	5.7578
Stddev	.1873
%RSD	3.2527

#1	6.0326
#2	5.5251
#3	5.8109
#4	5.7504
#5	5.6698

HNF-1663 REV.0

Method: ISO_U

Sample Name: iso-u std

Operator: bjg

Comment: 100 ppb

Run Time: 01/11/99 10:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.520	7.189	.7495	9.277	102.8
Stddev	.848	.338	.0493	.961	4.6
%RSD	15.36	4.696	6.580	10.36	4.520
#1	4.724	7.294	.6860	8.371	95.83
#2	6.063	7.240	.8241	10.86	106.2
#3	4.646	7.486	.7381	9.035	100.3
#4	6.608	6.608	.7494	9.391	105.2
#5	5.556	7.319	.7498	8.730	106.6

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	5.7578
Stddev	.1873
%RSD	3.2527

#1	6.0326
#2	5.5251
#3	5.8109
#4	5.7504
#5	5.6698

HNF-1663 REV. 0

Method: ISO_U Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/11/99 10:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_ [233]	U_ [234]	U_ [235]	U_ [236]	U_ [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.033	5.111	4.988	5.361	5.844
Stddev	.945	.507	.509	.572	.345
%RSD	18.77	9.928	10.20	10.67	5.909
#1	4.339	4.494	4.572	5.579	5.502
#2	4.003	5.809	5.338	4.945	6.358
#3	4.778	5.086	4.315	4.624	5.548
#4	6.066	5.366	5.444	5.599	5.910
#5	5.980	4.800	5.272	6.059	5.901

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	6.4189
Stddev	.0563
%RSD	.87704

#1	6.4526
#2	6.3695
#3	6.4885
#4	6.4293
#5	6.3544

Method: ISO_U Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/11/99 10:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.033	5.111	-.0032	5.361	.0012
Stddev	.945	.507	.0019	.572	.0014
%RSD	18.77	9.928	60.18	10.67	114.8
#1	4.339	4.494	-.0047	5.579	-.0002
#2	4.003	5.809	-.0019	4.945	.0033
#3	4.778	5.086	-.0057	4.624	.0000
#4	6.066	5.366	-.0015	5.599	.0015
#5	5.980	4.800	-.0021	6.059	.0014

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.4189
Stddev	.0563
%RSD	.87704

#1	6.4526
#2	6.3695
#3	6.4885
#4	6.4293
#5	6.3544

Method: ISO_U Sample Name: icv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/11/99 11:01 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.684	5.722	45.18	5.856	5072.
Stddev	.624	.414	2.97	.346	77.
%RSD	10.98	7.237	6.566	5.903	1.514
#1	6.015	5.558	47.97	5.863	5056.
#2	5.849	5.618	40.33	5.772	5086.
#3	6.375	5.453	44.62	5.991	5110.
#4	5.432	6.455	46.21	5.353	5155.
#5	4.747	5.525	46.77	6.303	4950.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.4701
Stddev	.0831
%RSD	1.2840

#1	6.5666
#2	6.4965
#3	6.5100
#4	6.3517
#5	6.4255

HNF-1663 REV. 0

Method: ISO_U Sample Name: icv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/11/99 11:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.684	5.722	.1471	5.856	20.38
Stddev	.624	.414	.0111	.346	.31
%RSD	10.98	7.237	7.539	5.903	1.516
#1	6.015	5.558	.1576	5.863	20.32
#2	5.849	5.618	.1290	5.772	20.43
#3	6.375	5.453	.1450	5.991	20.53
#4	5.432	6.455	.1510	5.353	20.71
#5	4.747	5.525	.1531	6.303	19.89

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.4701
Stddev	.0831
%RSD	1.2840

#1	6.5666
#2	6.4965
#3	6.5100
#4	6.3517
#5	6.4255

HNF-1663 REV. 0

Method: ISO_U

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/11/99 11:09 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.661	4.698	4.405	4.185	4.728
Stddev	.657	.772	.535	.542	.190
%RSD	14.09	16.43	12.15	12.96	4.008
#1	4.342	3.572	3.572	3.572	4.903
#2	4.631	4.631	4.269	4.052	4.776
#3	3.892	5.581	4.627	4.113	4.847
#4	4.771	4.478	4.551	5.065	4.698
#5	5.670	5.228	5.008	4.124	4.418

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.8919
Stddev	.1460
%RSD	2.1188

#1	7.1391
#2	6.9102
#3	6.8084
#4	6.8117
#5	6.7897

Method: ISO_U Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/11/99 11:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.661	4.698	- .0053	4.185	- .0033
Stddev	.657	.772	.0020	.542	.0008
%RSD	14.09	16.43	37.46	12.96	23.25
#1	4.342	3.572	- .0085	3.572	- .0026
#2	4.631	4.631	- .0059	4.052	- .0031
#3	3.892	5.581	- .0045	4.113	- .0028
#4	4.771	4.478	- .0048	5.065	- .0034
#5	5.670	5.228	- .0031	4.124	- .0045

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.8919
Stddev	.1460
%RSD	2.1188

#1	7.1391
#2	6.9102
#3	6.8084
#4	6.8117
#5	6.7897

Method: 'ISO_U Sample Name: s98t003011

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:11 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.737	5.078	5.449	4.280	7.956
Stddev	.483	.327	.702	.695	.911
%RSD	10.20	6.431	12.88	16.24	11.45
#1	5.428	5.127	4.825	4.900	9.122
#2	4.069	5.451	5.604	4.376	7.754
#3	4.776	5.324	5.638	4.933	8.222
#4	4.651	4.727	4.727	3.888	8.081
#5	4.760	4.760	6.449	3.301	6.603

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5203
Stddev	.0898
%RSD	1.3767

#1	6.6322
#2	6.5127
#3	6.3856
#4	6.5583
#5	6.5125

Method: ISO_U Sample Name: s98t003011

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.737	5.078	-.0014	4.280	.0097
Stddev	.483	.327	.0026	.695	.0037
%RSD	10.20	6.431	182.1	16.24	37.74
#1	5.428	5.127	-.0038	4.900	.0144
#2	4.069	5.451	-.0009	4.376	.0089
#3	4.776	5.324	-.0007	4.933	.0108
#4	4.651	4.727	-.0041	3.888	.0102
#5	4.760	4.760	.0023	3.301	.0043

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.5203
Stddev	.0898
%RSD	1.3767

#1	6.6322
#2	6.5127
#3	6.3856
#4	6.5583
#5	6.5125

Method: ISO_U Sample Name: s98t003011_d

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.870	5.101	5.063	5.065	8.405
Stddev	.488	.599	.440	.529	.710
%RSD	10.02	11.74	8.687	10.45	8.451
#1	4.871	5.096	5.620	4.646	8.318
#2	5.623	5.112	5.258	5.550	7.887
#3	4.392	4.464	4.968	4.968	7.631
#4	4.973	6.061	5.051	5.673	9.781
#5	4.489	4.774	4.418	4.489	9.406

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.7831
Stddev	.2339
%RSD	3.4488

#1	6.6725
#2	6.8470
#3	6.9450
#4	6.4342
#5	7.0169

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003011_d

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:13 Type: Unk

Mode: CONC

Corr-Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.870	5.101	-.0029	5.065	.0115
Stddev	.488	.599	.0016	.529	.0029
%RSD	10.02	11.74	57.03	10.45	24.82
#1	4.871	5.096	-.0008	4.646	.0112
#2	5.623	5.112	-.0022	5.550	.0094
#3	4.392	4.464	-.0032	4.968	.0084
#4	4.973	6.061	-.0029	5.673	.0130
#5	4.489	4.774	-.0053	4.489	.0155

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.7831
Stddev	.2339
%RSD	3.4488

#1	6.6725
#2	6.8470
#3	6.9450
#4	6.4342
#5	7.0169

INF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003012

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_ [233]	U_ [234]	U_ [235]	U_ [236]	U_ [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.874	4.997	4.673	4.548	8.322
Stddev	.637	1.147	.519	.536	.524
%RSD	13.06	22.95	11.12	11.80	6.292
#1	4.478	4.763	4.194	4.407	7.819
#2	5.464	6.362	5.015	4.491	8.009
#3	5.526	3.435	4.033	4.780	8.962
#4	4.861	5.848	5.165	3.796	8.810
#5	4.043	4.577	4.958	5.263	8.009

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	6.7095
Stddev	.1911
%RSD	2.8488

#1	7.0339
#2	6.6802
#3	6.6951
#4	6.5831
#5	6.5550

Method: ISO_U Sample Name: s98t003012

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.874	4.997	-.0043	4.548	.0112
Stddev	.637	1.147	.0019	.536	.0021
%RSD	13.06	22.95	44.73	11.80	18.84
#1	4.478	4.763	-.0061	4.407	.0092
#2	5.464	6.362	-.0031	4.491	.0099
#3	5.526	3.435	-.0067	4.780	.0138
#4	4.861	5.848	-.0025	3.798	.0131
#5	4.043	4.577	-.0033	5.263	.0099

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.7095
Stddev	.1911
%RSD	2.8488

#1	7.0339
#2	6.6802
#3	6.6951
#4	6.5831
#5	6.5550

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003012_d

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:18 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.021	4.585	4.898	4.428	7.538
Stddev	.427	.126	.413	.345	.957
%RSD	8.505	2.752	8.425	7.797	12.69
#1	5.559	4.620	4.620	4.404	7.869
#2	5.051	4.411	4.340	3.984	7.114
#3	4.774	4.625	5.371	4.849	8.952
#4	5.265	4.523	5.116	4.227	6.377
#5	4.456	4.748	5.040	4.675	7.378

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.8489
Stddev	.1331
%RSD	1.9440

#1	6.9261
#2	7.0285
#3	6.7024
#4	6.7429
#5	6.8447

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003012_d

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99

11:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	5.021	4.585	-.0035	4.428	.0080
Stddev	.427	.126	.0015	.345	.0038
%RSD	8.505	2.752	44.06	7.797	47.97
#1	5.559	4.620	-.0045	4.404	.0094
#2	5.051	4.411	-.0056	3.984	.0063
#3	4.774	4.625	-.0017	4.849	.0137
#4	5.265	4.523	-.0027	4.227	.0034
#5	4.456	4.748	-.0030	4.675	.0074

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.8489
Stddev	.1331
%RSD	1.9440

#1	6.9261
#2	7.0285
#3	6.7024
#4	6.7429
#5	6.8447

HNF-1663 REV. 0

Method: ISO_U

Sample Name: s98t003012_a

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:20 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.837	5.086	39.43	5.835	4599.
Stddev	.327	.624	2.16	.459	62.
%RSD	6.749	12.27	5.484	7.857	1.338
#1	4.721	5.021	41.59	6.070	4633.
#2	4.682	4.979	35.97	5.648	4554.
#3	4.769	5.796	40.06	6.163	4517.
#4	4.602	4.149	38.93	6.186	4665.
#5	5.410	5.484	40.61	5.114	4627.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	6.7180
Stddev	.0714
%RSD	1.0634

#1	6.6725
#2	6.7284
#3	6.8147
#4	6.6280
#5	6.7464

Method: ISO_U   Sample Name: s98t003012_a   Operator: bjg
Comment: 40501 df
Run Time: 01/11/99 11:20 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.837	5.086	.1256	5.836	18.48
Stddev	.327	.624	.0081	.459	.25
%RSD	6.749	12.27	6.437	7.857	1.339
#1	4.721	5.021	.1337	6.070	18.61
#2	4.682	4.979	.1127	5.648	18.29
#3	4.769	5.796	.1280	6.163	18.14
#4	4.602	4.149	.1237	6.186	18.74
#5	5.410	5.484	.1300	5.114	18.59

Int. Std.   Ir[193]
Line   193/pulse
Units   Cts/S
Avg   6.7180
Stddev   .0714
%RSD   1.0634

#1	6.6725
#2	6.7284
#3	6.8147
#4	6.6280
#5	6.7464

Method: ISO_U

Sample Name: s98t003013

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99 11:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.177	4.247	4.434	4.110	7.612
Stddev	.674	.646	.793	.662	.623
%RSD	16.13	15.21	17.89	16.10	8.187
#1	3.432	3.310	4.107	3.187	7.110
#2	4.354	5.045	5.528	4.423	6.841
#3	4.785	4.142	4.499	3.642	8.213
#4	4.814	4.608	4.677	4.677	7.703
#5	3.502	4.132	3.362	4.622	8.194

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.3608
Stddev	.4574
%RSD	6.2143

#1	8.1576
#2	7.2353
#3	7.0014
#4	7.2703
#5	7.1394

Method: ISO_U   Sample Name: s98t003013   Operator: bjg
Comment: 40501 df
Run Time: 01/11/99 11:27 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.177	4.247	-.0052	4.110	.0083
Stddev	.674	.646	.0030	.662	.0025
%RSD	16.13	15.21	56.69	16.10	30.12

#1	3.432	3.310	-.0065	3.187	.0063
#2	4.354	5.045	-.0011	4.423	.0052
#3	4.785	4.142	-.0050	3.642	.0107
#4	4.814	4.608	-.0043	4.677	.0087
#5	3.502	4.132	-.0092	4.622	.0107

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.3608
Stddev	.4574
%RSD	6.2143

#1	8.1576
#2	7.2353
#3	7.0014
#4	7.2703
#5	7.1394

Method: ISO_U

Sample Name: s98t003013_d

Operator: bjg

Comment: 40501 df

Run Time: 01/11/99

11:28

Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.073	4.267	4.494	4.514	7.961
Stddev	.243	.419	.528	.467	.205
%RSD	5.977	9.810	11.74	10.34	2.576
#1	3.847	4.967	5.247	3.917	7.975
#2	3.985	4.335	4.055	4.195	7.830
#3	3.868	3.940	3.940	5.086	8.310
#4	4.373	4.086	4.516	4.803	7.814
#5	4.289	4.008	4.711	4.570	7.875

Int. Std.

Ir[193]

Line	193/pulse
Units	Cts/S
Avg	7.0730
Stddev	.0888
%RSD	1.2558

#1	7.1476
#2	7.1518
#3	6.9795
#4	6.9750
#5	7.1110

Method: ISO_U

Sample Name: s98t003013_d

Operator: bjg

Comment: 40501.d

Run Time: 01/11/99

11:28

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.073	4.267	-.0050	4.514	.0097
Stddev	.243	.419	.0020	.467	.0008
%RSD	5.977	9.810	39.36	10.34	8.483
#1	3.847	4.967	-.0022	3.917	.0098
#2	3.985	4.335	-.0067	4.195	.0092
#3	3.868	3.940	-.0071	5.086	.0111
#4	4.373	4.086	-.0049	4.803	.0091
#5	4.289	4.008	-.0042	4.570	.0094

Int. Std.

Ir[193]

Line	193/pulse
Units	Cts/S
Avg	7.0730
Stddev	.0888
%RSD	1.2558
#1	7.1476
#2	7.1518
#3	6.9795
#4	6.9750
#5	7.1110

Method: ISO_U Sample Name: ccv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/11/99 11:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.623	4.829	43.55	5.404	5009.
Stddev	.692	.758	1.73	.786	139.
%RSD	14.97	15.70	3.980	14.54	2.781
#1	4.452	4.587	41.69	6.206	4862.
#2	4.193	3.774	45.42	4.822	5055.
#3	3.871	4.957	44.82	5.772	4944.
#4	4.966	5.876	41.76	5.876	5226.
#5	5.631	4.953	44.03	4.342	4959.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.2894
Stddev	.1273
%RSD	1.7456

#1	7.4116
#2	7.1551
#3	7.3627
#4	7.1480
#5	7.3695

Analysis Report

HNF-1663 REV. 0

Method: ISO_U Sample Name: ccv

Operator: bjg

Comment: 20 ppb u

Run Time: 01/11/99 11:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.623	4.829	.1410	5.404	20.13
Stddev	.692	.758	.0065	.786	.56
%RSD	14.97	15.70	4.595	14.54	2.784
#1	4.452	4.587	.1341	6.206	19.53
#2	4.193	3.774	.1480	4.822	20.31
#3	3.871	4.957	.1458	5.772	19.86
#4	4.966	5.876	.1343	5.876	21.00
#5	5.631	4.953	.1428	4.342	19.92

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	7.2894
Stddev	.1273
%RSD	1.7456

#1	7.4116
#2	7.1551
#3	7.3627
#4	7.1480
#5	7.3695

Method: ISO_U Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/11/99 11:35 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.268	4.160	4.517	4.053	4.457
Stddev	.780	.437	.229	.516	.333
%RSD	18.27	10.51	5.064	12.72	7.477

#1	3.157	4.129	4.554	3.339	4.371
#2	4.335	4.203	4.203	4.335	4.729
#3	4.373	4.695	4.437	4.695	4.502
#4	5.346	4.290	4.554	4.092	4.752
#5	4.127	3.482	4.836	3.804	3.933

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.7905
Stddev	.2632
%RSD	3.3779

#1	8.2352
#2	7.6128
#3	7.7747
#4	7.5759
#5	7.7540

Analysis Report

HNF-1663 REV. 0

Method: ISO_U Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/11/99 11:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	4.268	4.160	-.0049	4.053	-.0044
Stddev	.780	.437	.0009	.516	.0013
%RSD	18.27	10.51	17.36	12.72	30.70
#1	3.157	4.129	-.0048	3.339	-.0047
#2	4.335	4.203	-.0061	4.335	-.0033
#3	4.373	4.695	-.0052	4.695	-.0042
#4	5.346	4.290	-.0048	4.092	-.0032
#5	4.127	3.482	-.0037	3.804	-.0065

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	7.7905
Stddev	.2632
%RSD	3.3779

#1	8.2352
#2	7.6128
#3	7.7747
#4	7.5759
#5	7.7540

LABCORE Completed Worklist Report for Worklist# 27507

Analyst: kjt

Instrument: ICPMS1

Book#: _____

Method: LA-506-101 Rev/Mod _____

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSPR

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		QMSPR-QC TC99	QC	0.0197	1.70e-02	86.294 %	Recovery
1 ICV	0		QMSPR-QC SN117	QC	0.0077	6.46e-03	83.896 %	Recovery
1 ICV	0		QMSPR-QC SN126	QC	n/a	<1.50e-4		% Recovery
1 ICV	0		QMSPR-QC CS133	QC	0.0200	2.19e-02	109.500 %	Recovery
1 ICV	0		QMSPR-QC CS135	QC	n/a	4.52e-05		% Recovery
1 ICV	0		QMSPR-QC CS137	QC	n/a	7.66e-05		% Recovery
2 ICB	0		QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
2 ICB	0		QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
2 ICB	0		QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
2 ICB	0		QMSPR-QC CS133	QC	1	1.04e-05	1.040e-005	ug/mL
2 ICB	0		QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
2 ICB	0		QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL
3 BLANK-PREP	0		QMSPR-A1 TC99-A1	SOLID	1	3.30e-05	3.300e-005	ug/g
3 BLANK-PREP	0		QMSPR-A1 SN117-A1	SOLID	1	6.82e-03	0.007	ug/g
3 BLANK-PREP	0		QMSPR-A1 SN126-A1	SOLID	1	<1.50e-3		ug/g
3 BLANK-PREP	0		QMSPR-A1 CS133-A1	SOLID	1	9.90e-05	9.900e-005	ug/g
3 BLANK-PREP	0		QMSPR-A1 CS135-A1	SOLID	1	3.90e-05	3.900e-005	ug/g
3 BLANK-PREP	0		QMSPR-A1 CS137-A1	SOLID	1	5.30e-05	5.300e-005	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	1.173e+00	7.80e-002	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	< 5.848e+00	5.847	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	< 5.848e+00	5.847	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	2.748e+00	7.80e-002	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	1.867e+00	7.80e-002	ug/g
4 SAMPLE	S98T002990	0 A	QMSPR-A1 CS137-A1	SOLID	N/A	3.060e+00	7.80e-002	ug/g
5 DUP	S98T002990	0 A	QMSPR-A1 TC99-A1	SOLID	1.17e+00	1.08e+00	8.000	RPD
5 DUP	S98T002990	0 A	QMSPR-A1 SN117-A1	SOLID	<5.85e0	<6.00e0		RPD
5 DUP	S98T002990	0 A	QMSPR-A1 SN126-A1	SOLID	<5.85e0	<6.00e0		RPD
5 DUP	S98T002990	0 A	QMSPR-A1 CS133-A1	SOLID	2.75e+00	2.74e+00	0.364	RPD
5 DUP	S98T002990	0 A	QMSPR-A1 CS135-A1	SOLID	1.87e+00	2.15e+00	13.930	RPD
5 DUP	S98T002990	0 A	QMSPR-A1 CS137-A1	SOLID	3.06e+00	3.36e+00	9.346	RPD
6 SPK-POST	S98T002990	0 A	QMSPR-A1 TC99-A1	SOLID	0.0219	1.85e-02	84.475 %	Recovery
6 SPK-POST	S98T002990	0 A	QMSPR-A1 SN117-A1	SOLID	n/a	4.10e-03		% Recovery
6 SPK-POST	S98T002990	0 A	QMSPR-A1 SN126-A1	SOLID	n/a	5.00e-07		% Recovery
6 SPK-POST	S98T002990	0 A	QMSPR-A1 CS133-A1	SOLID	0.0201	2.23e-02	110.945 %	Recovery
6 SPK-POST	S98T002990	0 A	QMSPR-A1 CS135-A1	SOLID	n/a	-1.10e-6		% Recovery
6 SPK-POST	S98T002990	0 A	QMSPR-A1 CS137-A1	SOLID	n/a	-3.70e-6		% Recovery
7 SAMPLE	S98T002991	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	1.211e+00	7.97e-002	ug/g
7 SAMPLE	S98T002991	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	< 5.977e+00	5.978	ug/g
7 SAMPLE	S98T002991	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	< 5.977e+00	5.978	ug/g
7 SAMPLE	S98T002991	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	2.503e+00	7.97e-002	ug/g
7 SAMPLE	S98T002991	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	1.494e+00	7.97e-002	ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27507

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
7 SAMPLE	S98T002991	0 A	QMSPR-A1 CS137-A1 SOLID	<u>N/A</u>	2.423e+00	7.97e-002	ug/g
8 DUP	S98T002991	0 A	QMSPR-A1 TC99-A1 SOLID	1.21e+00	1.14e+00	5.957	RPD
8 DUP	S98T002991	0 A	QMSPR-A1 SN117-A1 SOLID	<5.98e0	<6.12e0		RPD
8 DUP	S98T002991	0 A	QMSPR-A1 SN126-A1 SOLID	<5.98e0	<6.12e0		RPD
8 DUP	S98T002991	0 A	QMSPR-A1 CS133-A1 SOLID	2.50e+00	2.25e+00	10.526	RPD
8 DUP	S98T002991	0 A	QMSPR-A1 CS135-A1 SOLID	1.49e+00	1.48e+00	0.673	RPD
8 DDP	S98T002991	0 A	QMSPR-A1 CS137-A1 SOLID	2.42e+00	2.45e+00	1.232	RPD
9 SPK-POST	S98T002991	0 A	QMSPR-A1 TC99-A1 SOLID	0.0219	1.88e-02	85.845	% Recovery
9 SPK-POST	S98T002991	0 A	QMSPR-A1 SN117-A1 SOLID	n/a	4.08e-03		% Recovery
9 SPK-POST	S98T002991	0 A	QMSPR-A1 SN126-A1 SOLID	n/a	1.00e-06		% Recovery
9 SPK-POST	S98T002991	0 A	QMSPR-A1 CS133-A1 SOLID	0.0201	2.28e-02	113.433	% Recovery
9 SPK-POST	S98T002991	0 A	QMSPR-A1 CS135-A1 SOLID	n/a	1.70e-06		% Recovery
9 SPK-POST	S98T002991	0 A	QMSPR-A1 CS137-A1 SOLID	n/a	3.70e-06		% Recovery
10 SAMPLE	S98T002992	0 A	QMSPR-A1 TC99-A1 SOLID	<u>N/A</u>	5.302e-01	7.85e-002	ug/g
10 SAMPLE	S98T002992	0 A	QMSPR-A1 SN117-A1 SOLID	<u>N/A</u> <	5.891e+00	5.891	ug/g
10 SAMPLE	S98T002992	0 A	QMSPR-A1 SN126-A1 SOLID	<u>N/A</u> <	5.891e+00	5.891	ug/g
10 SAMPLE	S98T002992	0 A	QMSPR-A1 CS133-A1 SOLID	<u>N/A</u>	1.936e+00	7.85e-002	ug/g
10 SAMPLE	S98T002992	0 A	QMSPR-A1 CS135-A1 SOLID	<u>N/A</u>	1.335e+00	7.85e-002	ug/g
10 SAMPLE	S98T002992	0 A	QMSPR-A1 CS137-A1 SOLID	<u>N/A</u>	2.635e+00	7.85e-002	ug/g
11 DUP	S98T002992	0 A	QMSPR-A1 TC99-A1 SOLID	5.30e-01	6.10e-01	14.035	RPD
11 DUP	S98T002992	0 A	QMSPR-A1 SN117-A1 SOLID	<5.89e0	<5.98e0		RPD
11 DUP	S98T002992	0 A	QMSPR-A1 SN126-A1 SOLID	<5.89e0	<5.98e0		RPD
11 DUP	S98T002992	0 A	QMSPR-A1 CS133-A1 SOLID	1.94e+00	1.68e+00	14.365	RPD
11 DUP	S98T002992	0 A	QMSPR-A1 CS135-A1 SOLID	1.33e+00	1.51e+00	12.676	RPD
11 DUP	S98T002992	0 A	QMSPR-A1 CS137-A1 SOLID	2.63e+00	2.69e+00	2.256	RPD
12 SAMPLE	S98T002993	0 A	QMSPR-A1 TC99-A1 SOLID	<u>N/A</u>	1.010e+00	8.02e-002	ug/g
12 SAMPLE	S98T002993	0 A	QMSPR-A1 SN117-A1 SOLID	<u>N/A</u> <	6.013e+00	6.014	ug/g
12 SAMPLE	S98T002993	0 A	QMSPR-A1 SN126-A1 SOLID	<u>N/A</u> <	6.013e+00	6.014	ug/g
12 SAMPLE	S98T002993	0 A	QMSPR-A1 CS133-A1 SOLID	<u>N/A</u>	1.816e+00	8.02e-002	ug/g
12 SAMPLE	S98T002993	0 A	QMSPR-A1 CS135-A1 SOLID	<u>N/A</u>	1.038e+00	8.02e-002	ug/g
12 SAMPLE	S98T002993	0 A	QMSPR-A1 CS137-A1 SOLID	<u>N/A</u>	1.567e+00	8.02e-002	ug/g
13 DUP	S98T002993	0 A	QMSPR-A1 TC99-A1 SOLID	1.01e+00	1.14e+00	12.093	RPD
13 DUP	S98T002993	0 A	QMSPR-A1 SN117-A1 SOLID	<6.01e0	<6.19e0		RPD
13 DUP	S98T002993	0 A	QMSPR-A1 SN126-A1 SOLID	<6.01e0	<6.19e0		RPD
13 DUP	S98T002993	0 A	QMSPR-A1 CS133-A1 SOLID	1.82e+00	1.92e+00	5.348	RPD
13 DUP	S98T002993	0 A	QMSPR-A1 CS135-A1 SOLID	1.04e+00	1.11e+00	6.512	RPD
13 DUP	S98T002993	0 A	QMSPR-A1 CS137-A1 SOLID	1.57e+00	1.66e+00	5.573	RPD
14 CCV	0	QMSPR-QC TC99	QC	0.0197	1.73e-02	87.817	% Recovery
14 CCV	0	QMSPR-QC SN117	QC	0.0077	6.43e-03	83.506	% Recovery
14 CCV	0	QMSPR-QC SN126	QC	n/a	<1.50e-4		% Recovery
14 CCV	0	QMSPR-QC CS133	QC	0.0200	2.19e-02	109.500	% Recovery
14 CCV	0	QMSPR-QC CS135	QC	n/a	4.60e-05		% Recovery
14 CCV	0	QMSPR-QC CS137	QC	n/a	7.85e-05		% Recovery
15 CCB	0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
15 CCB	0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
15 CCB	0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
15 CCB	0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
15 CCB	0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
15 CCB	0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27507

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Final page for worklist# 27507

_____
Analyst Signature   Date

_____
Analyst Signature   Date

_____
Reviewer Signature   Date

12/10/98 06:45
A-0004-1

FILE # PRO125A-TAT

Page: 1

LABCORE Data Entry Template for Worklist# 27507

Analyst: LT Instrument: ICPMS1 Book# 42857 980125-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST) MSPR

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSPR-QC	QC		
2	ICB		@MSPR-QC	QC		
3	BLNK-PREP		@MSPR-A1	SOLID		
4	SAMPLE	S98T002990 0 A	@MSPR-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
5	DUP	S98T002990 0 A	@MSPR-A1	SOLID		
6	SPK-POST	S98T002990 0 A	@MSPR-A1	SOLID		
7	SAMPLE	S98T002991 0 A	@MSPR-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
8	DUP	S98T002991 0 A	@MSPR-A1	SOLID		
9	SPK-POST	S98T002991 0 A	@MSPR-A1	SOLID		
10	SAMPLE	S98T002992 0 A	@MSPR-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
11	DUP	S98T002992 0 A	@MSPR-A1	SOLID		
12	SAMPLE	S98T002993 0 A	@MSPR-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
13	DUP	S98T002993 0 A	@MSPR-A1	SOLID		
14	CCV		@MSPR-QC	QC		
15	CCB		@MSPR-QC	QC		

Data Entry Comments:

Uploaded + Validated by:

Paul H. Pang

1/25/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2587

LABCORE Data Entry Template for Worklist# 27507

S Type	Sample#	R A Test	Matrix	Group#	Project

Final page for worklist # 27507

Office
Analyst Signature Date 1-25-99

Analyst Signature Date

BLK - 1144 1 ml - 8-1
598700 2990 05-10-1

DF
10
201

2990-D
2990-A
2991
2991-D
2991-A
2992
2992-D
2993
2993-D

POST SPIKES = 1 ml of 44 B57

SCANS = PR 2990, PR 2991, PR 2992, PR 2993

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2588

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 01/25/99 10:08 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	88210.
Stddev	1062.
%RSD	1.204

#1	89140.
#2	88380.
#3	88750.
#4	86310.
#5	89180.
#6	86780.
#7	87880.
#8	88930.
#9	89310.
#10	87480.

[Signature]
1-25-99

B4-105 AND BEST

528500 2990, 2991, 2992, 2993

05-10-1

FILED PRO1257.787

W.C. # 27807

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2589 TO 2610

Method: TC_SN_CS Sample Name: Blank

Operator:

Comment:

Run Time: 01/25/99 10:10 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.116	1.631	6.661	2.548	1.041
Stddev	.402	.201	1.037	.331	.336
%RSD	18.99	12.31	15.57	12.99	32.25

#1	1.558	1.947	5.897	2.114	.7788
#2	2.528	1.436	6.146	2.413	1.321
#3	2.010	1.564	5.808	3.016	1.396
#4	1.995	1.511	7.253	2.538	1.088
#5	2.489	1.697	8.202	2.659	.6222

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	8.259	4.633	6.713	32.20
Stddev	.469	.325	.320	1.04
%RSD	5.678	7.015	4.764	3.241

#1	7.566	4.395	6.342	31.88
#2	8.674	4.940	6.491	32.23
#3	8.265	4.189	6.925	32.45
#4	8.703	4.775	7.132	33.67
#5	8.089	4.865	6.675	30.77

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.7515
Stddev	.2898
%RSD	3.3113

#1	8.9881
#2	8.7042
#3	8.9530
#4	8.2726
#5	8.8395

Method: TC_SN_CS Sample Name: PRIV Std

Operator:

Comment:

Run Time: 01/25/99 10:15 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]
Line	99/pulse	100/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	13800.	14160.	8205.	8205.
Stddev	265.	263.	211.	211.
%RSD	1.921	1.855	2.569	2.569

#1	13840.	14430.	7986.	7986.
#2	14240.	14350.	8545.	8545.
#3	13770.	13780.	8235.	8235.
#4	13580.	14030.	8100.	8100.
#5	13590.	14220.	8159.	8159.

Elem	Te[125]	Cs[133]	Cs[135]	Cs[137]
Line	125/pulse	133/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1640.	87780.	6375.	10950.
Stddev	73.	2677.	241.	334.
%RSD	4.444	3.050	3.785	3.047

#1	1737.	85370.	6400.	11140.
#2	1697.	91710.	6776.	11430.
#3	1607.	88870.	6184.	10710.
#4	1581.	85290.	6319.	10610.
#5	1578.	87640.	6196.	10880.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	66970.
Stddev	2766.
%RSD	4.130

#1	65410.
#2	71720.
#3	65290.
#4	65310.
#5	67110.

Sample Name: PRIV Std Run Time: 01/25/99 10:15

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.2459
Stddev	.1698
%RSD	2.0596

#1	8.1908
#2	8.0120
#3	8.2033
#4	8.4318
#5	8.3916

Method: TC_SN_CS Sample Name: blank

Operator: bjg

Comment: background

Run Time: 01/25/99 10:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	.0002	.0010	.0002	-.0005
Stddev	.0005	.0004	.0005	.0003	.0008
%RSD	285.7	221.6	53.30	141.5	149.6
#1	.0006	.0005	.0012	.0005	-.0003
#2	.0008	-.0005	.0006	.0000	-.0015
#3	-.0003	.0003	.0004	.0003	-.0003
#4	.0001	.0002	.0018	.0003	.0005
#5	-.0004	.0004	.0012	-.0002	-.0009

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0114	.0000	-.0003	.0002
Stddev	.0009	.001	.0004	.0002
%RSD	7.661	2653.	116.4	96.12
#1	.0116	-.0006	-.0002	.0003
#2	.0126	.0006	-.0002	.0000
#3	.0118	.0013	-.0001	.0006
#4	.0104	-.0008	-.0009	.0001
#5	.0108	-.0006	-.0001	.0002

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.8704
Stddev	.2176
%RSD	2.4527

#1	8.6863
#2	8.7485
#3	8.7027
#4	9.1159
#5	9.0989

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/25/99 10:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0039	.0003	6.456	.0001	.0002
Stddev	.0003	.0002	.065	.0006	.0011
%RSD	8.617	58.04	1.006	1016.	484.7
#1	.0036	.0004	6.375	-.0002	-.0004
#2	.0038	.0001	6.509	.0009	.0001
#3	.0043	.0006	6.535	-.0007	.0021
#4	.0035	.0005	6.422	.0002	-.0004
#5	.0041	.0002	6.442	.0001	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0153	-.0007	-.0031	-.0019
Stddev	.0021	.0007	.0004	.0003
%RSD	13.63	94.25	12.17	15.35
#1	.0167	-.0005	-.0034	-.0020
#2	.0179	-.0008	-.0035	-.0019
#3	.0155	-.0019	-.0029	-.0020
#4	.0130	.0000	-.0027	-.0022
#5	.0135	-.0005	-.0028	-.0014

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.5162
Stddev	.0706
%RSD	.82937

#1	8.4993
#2	8.4502
#3	8.5677
#4	8.4539
#5	8.6099

Method: TC_SN_CS Sample Name: icv

Operator: bjg

Comment: msprv. lcs

Run Time: 01/25/99 10:34 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.02	.0076	1.421	.0002	.0008
Stddev	.39	.0009	.053	.0005	.0011
%RSD	2.319	11.53	3.746	283.7	137.2

#1	17.56	.0065	1.479	.0009	.0013
#2	16.68	.0079	1.394	.0002	-.0008
#3	17.31	.0074	1.479	-.0003	.0019
#4	16.72	.0072	1.367	-.0002	.0016
#5	16.83	.0089	1.389	.0002	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.90	.0452	.0766	.0523
Stddev	.45	.0027	.0035	.0011
%RSD	2.052	5.903	4.608	2.196

#1	22.44	.0451	.0804	.0522
#2	22.06	.0497	.0793	.0519
#3	21.30	.0428	.0714	.0528
#4	22.10	.0440	.0762	.0538
#5	21.60	.0443	.0757	.0507

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.7035
Stddev	.1206
%RSD	1.3859

#1	8.5832
#2	8.8290
#3	8.7682
#4	8.5652
#5	8.7718

Method: TC_SN_CS Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 01/25/99 10:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	.0002	.0004	.0001	-.0004
Stddev	.0006	.0003	.0007	.0005	.0012
%RSD	161.9	139.1	171.9	436.9	323.8

#1	.0012	.0000	.0004	.0000	-.0022
#2	-.0003	.0002	.0003	.0005	.0010
#3	-.0001	.0006	-.0006	.0001	.0004
#4	.0004	.0003	.0005	-.0006	-.0006
#5	.0005	-.0001	.0013	.0006	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0104	-.0003	-.0003	.0000
Stddev	.0008	.0005	.0008	.000
%RSD	7.804	146.3	310.3	629.0

#1	.0089	-.0007	-.0009	-.0003
#2	.0107	.0002	-.0009	.0000
#3	.0108	-.0008	.0004	.0000
#4	.0107	-.0005	-.0007	.0000
#5	.0107	.0002	.0008	.0001

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.9427
Stddev	.2520
%RSD	2.8175

#1	9.3477
#2	9.0193
#3	8.7619
#4	8.8471
#5	8.7373

Method: TC_SN_CS Sample Name: blnk-prep

Operator: bjg

Comment: 10 df

Run Time: 01/25/99 10:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0033	.0017	.6820	.0004	.0007
Stddev	.0005	.0003	.0057	.0005	.0011
%RSD	16.10	15.12	.8398	117.4	144.5
#1	.0030	.0015	.6809	.0003	.0002
#2	.0038	.0016	.6802	.0001	.0010
#3	.0029	.0016	.6770	.0000	.0017
#4	.0029	.0021	.6801	.0011	.0016
#5	.0039	.0019	.6919	.0007	.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0099	.0039	.0053	.0037
Stddev	.0012	.0006	.0006	.0002
%RSD	11.95	15.74	10.78	5.036
#1	.0107	.0038	.0049	.0035
#2	.0078	.0033	.0054	.0037
#3	.0103	.0035	.0060	.0039
#4	.0103	.0039	.0045	.0035
#5	.0102	.0049	.0057	.0037

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.0072
Stddev	.1400
%RSD	1.5546

#1	9.1888
#2	9.0468
#3	9.0583
#4	8.9167
#5	8.8253

Method: TC_SN_CS Sample Name: s98t002990

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 10:51 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0301	.0369	.0266	.0004	-.0004
Stddev	.0018	.0011	.0016	.0006	.0012
%RSD	5.904	3.016	6.145	143.0	307.8

#1	.0292	.0356	.0281	.0003	-.0006
#2	.0314	.0380	.0279	-.0002	-.0009
#3	.0311	.0369	.0265	-.0001	.0006
#4	.0315	.0380	.0240	.0011	.0010
#5	.0274	.0359	.0265	.0009	-.0019

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0705	.0479	.0785	.0293
Stddev	.0042	.0025	.0021	.0009
%RSD	6.017	5.243	2.691	3.140

#1	.0641	.0443	.0772	.0281
#2	.0682	.0476	.0754	.0290
#3	.0731	.0474	.0795	.0292
#4	.0743	.0492	.0805	.0296
#5	.0727	.0511	.0798	.0307

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.5243
Stddev	.2314
%RSD	2.7151

#1	8.9110
#2	8.4988
#3	8.4687
#4	8.2869
#5	8.4564

Method: TC_SN_CS

Sample Name: s98t002990_d

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 10:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0270	.0362	.0443	.0001	-.0010
Stddev	.0027	.0009	.0021	.0003	.0010
%RSD	10.04	2.543	4.757	514.6	97.71
#1	.0314	.0377	.0467	.0003	-.0010
#2	.0250	.0364	.0463	.0001	-.0013
#3	.0248	.0358	.0428	-.0004	-.0016
#4	.0260	.0353	.0435	-.0001	.0007
#5	.0278	.0357	.0420	.0004	-.0017

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0685	.0537	.0839	.0365
Stddev	.0050	.0026	.0043	.0008
%RSD	7.240	4.891	5.154	2.264
#1	.0727	.0562	.0853	.0375
#2	.0670	.0542	.0817	.0361
#3	.0668	.0495	.0799	.0354
#4	.0618	.0531	.0817	.0368
#5	.0740	.0555	.0908	.0368

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.7727
Stddev	.1602
%RSD	1.8262

#1	8.7447
#2	8.7704
#3	9.0188
#4	8.7593
#5	8.5699

2599

Method: TC_SN_CS Sample Name: s98t002990_a

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 10:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.51	.0377	4.096	.0005	.0001
Stddev	1.02	.0009	.136	.0003	.0010
%RSD	5.510	2.432	3.314	51.19	1695.
#1	19.90	.0384	4.261	.0004	-.0004
#2	17.19	.0368	4.140	.0005	-.0012
#3	18.40	.0374	4.020	.0004	.0013
#4	18.04	.0388	3.908	.0010	.0007
#5	19.00	.0369	4.154	.0004	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	22.36	.0468	.0748	.0295
Stddev	.96	.0039	.0044	.0013
%RSD	4.308	8.389	5.913	4.421
#1	23.18	.0506	.0798	.0303
#2	20.75	.0461	.0770	.0290
#3	22.71	.0467	.0716	.0295
#4	22.24	.0407	.0689	.0276
#5	22.90	.0499	.0767	.0310

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.1588
Stddev	.1739
%RSD	2.1308

#1	7.9956
#2	8.3854
#3	8.1315
#4	8.2848
#5	7.9967

2600

Method: TC_SN_CS Sample Name: s98t002991

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 11:08 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0304	.0349	.0299	.0003	.0005
Stddev	.0016	.0014	.0016	.0003	.0007
%RSD	5.147	3.961	5.189	117.4	143.5

#1	.0308	.0349	.0277	.0001	.0004
#2	.0295	.0361	.0305	.0007	.0004
#3	.0286	.0359	.0308	-.0001	.0007
#4	.0328	.0351	.0315	.0006	.0013
#5	.0303	.0326	.0289	.0003	-.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0628	.0375	.0608	.0195
Stddev	.0035	.0025	.0031	.0014
%RSD	5.563	6.558	5.081	7.087

#1	.0617	.0417	.0579	.0181
#2	.0624	.0364	.0602	.0209
#3	.0625	.0375	.0643	.0191
#4	.0684	.0366	.0637	.0211
#5	.0588	.0354	.0578	.0184

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.2968
Stddev	.2445
%RSD	2.9469

#1	8.3905
#2	7.9804
#3	8.3553
#4	8.1409
#5	8.6169

2601

Method: TC_SN_CS

Sample Name: s98t002991_d

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 11:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0280	.0341	.0540	.0005	-.0010
Stddev	.0016	.0012	.0021	.0006	.0009
%RSD	5.569	3.529	3.880	105.8	88.17

#1	.0262	.0337	.0545	.0002	-.0017
#2	.0283	.0352	.0506	.0015	.0004
#3	.0284	.0344	.0560	.0005	-.0014
#4	.0267	.0349	.0553	.0001	-.0018
#5	.0302	.0322	.0536	.0004	-.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0550	.0362	.0600	.0210
Stddev	.0038	.0024	.0045	.0016
%RSD	6.916	6.738	7.552	7.650

#1	.0561	.0394	.0671	.0229
#2	.0602	.0380	.0608	.0222
#3	.0560	.0344	.0593	.0207
#4	.0511	.0353	.0584	.0201
#5	.0515	.0338	.0546	.0189

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.5503
Stddev	.3010
%RSD	3.5201

#1	8.3593
#2	8.3832
#3	8.3289
#4	8.6388
#5	9.0413

Method: TC_SN_CS

Sample Name: s98t002991_a

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 11:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.88	.0382	4.078	.0010	.0003
Stddev	.61	.0016	.230	.0002	.0016
%RSD	3.220	4.268	5.630	23.74	542.8

#1	19.23	.0401	3.826	.0008	-.0012
#2	17.80	.0366	4.183	.0013	-.0009
#3	19.23	.0396	4.234	.0009	-.0004
#4	19.02	.0366	3.838	.0007	.0016
#5	19.09	.0382	4.311	.0012	.0023

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	22.91	.0392	.0645	.0220
Stddev	.84	.0017	.0039	.0012
%RSD	3.667	4.340	5.986	5.545

#1	23.32	.0408	.0696	.0214
#2	21.87	.0389	.0638	.0231
#3	22.79	.0369	.0639	.0220
#4	22.49	.0386	.0590	.0203
#5	24.08	.0410	.0661	.0231

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	7.8509
Stddev	.1846
%RSD	2.3515

#1	7.8469
#2	7.7623
#3	7.8712
#4	8.1374
#5	7.6366

2603

Method: TC_SN_CS Sample Name: s98t002992

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 11:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0135	.0173	.0238	.0003	-.0007
Stddev	.0011	.0015	.0021	.0002	.0006
%RSD	8.274	8.538	8.821	92.75	83.78

#1	.0150	.0189	.0254	.0004	.0000
#2	.0143	.0153	.0221	.0006	-.0010
#3	.0131	.0167	.0245	-.0001	-.0002
#4	.0132	.0169	.0212	.0003	-.0014
#5	.0121	.0185	.0260	.0002	-.0008

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0493	.0340	.0671	.0242
Stddev	.0019	.0025	.0056	.0011
%RSD	3.923	7.416	8.358	4.377

#1	.0520	.0357	.0759	.0261
#2	.0473	.0323	.0659	.0238
#3	.0486	.0315	.0676	.0239
#4	.0505	.0331	.0655	.0234
#5	.0481	.0376	.0604	.0239

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.2534
Stddev	.3606
%RSD	4.3692

#1	7.7034
#2	8.5463
#3	8.5707
#4	8.3445
#5	8.1021

Method: TC_SN_CS Sample Name: s98t002992_d Operator: bjg
Comment: 201 df
Run Time: 01/25/99 11:27 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0153	.0165	.0356	.0001	.0001
Stddev	.0008	.0011	.0023	.0003	.0008
%RSD	5.556	6.635	6.374	292.7	668.2

#1	.0161	.0151	.0339	-.0002	-.0002
#2	.0141	.0158	.0334	.0000	-.0011
#3	.0158	.0179	.0383	.0005	.0006
#4	.0158	.0171	.0378	.0004	.0006
#5	.0147	.0167	.0347	-.0002	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0422	.0379	.0674	.0281
Stddev	.0043	.0023	.0056	.0011
%RSD	10.14	6.164	8.242	4.012

#1	.0379	.0342	.0611	.0274
#2	.0455	.0370	.0629	.0270
#3	.0467	.0395	.0735	.0294
#4	.0434	.0396	.0724	.0292
#5	.0375	.0393	.0669	.0273

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.2522
Stddev	.2407
%RSD	2.9170

#1	8.5421
#2	8.2940
#3	8.0230
#4	7.9910
#5	8.4110

Method: TC_SN_CS Sample Name: s98t002993

Operator: bjg

Comment 201 df

Run Time: 01/25/99 11:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0252	.0291	.0217	.0003	-.0004
Stddev	.0015	.0021	.0010	.0005	.0007
%RSD	6.013	7.105	4.723	213.2	181.9

#1	.0237	.0303	.0225	.0005	-.0002
#2	.0272	.0312	.0228	-.0003	-.0001
#3	.0238	.0261	.0212	-.0003	-.0015
#4	.0251	.0278	.0219	.0004	.0004
#5	.0261	.0300	.0203	.0009	-.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0453	.0259	.0391	.0108
Stddev	.0009	.0026	.0015	.0004
%RSD	2.002	9.861	3.820	3.536

#1	.0459	.0245	.0370	.0102
#2	.0443	.0293	.0409	.0108
#3	.0445	.0275	.0384	.0110
#4	.0451	.0252	.0394	.0112
#5	.0465	.0228	.0398	.0108

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.3209
Stddev	.2899
%RSD	3.4839

#1	8.4981
#2	7.8160
#3	8.5123
#4	8.3449
#5	8.4333

Method: TC_SN_CS Sample Name: s98t002993_d

Operator: bjg

Comment: 201 df

Run Time: 01/25/99 11:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0277	.0323	.0489	.0002	-.0004
Stddev	.0015	.0019	.0023	.0006	.0008
%RSD	5.486	6.021	4.696	277.0	195.2
#1	.0270	.0350	.0477	.0002	.0002
#2	.0258	.0330	.0478	-.0004	-.0016
#3	.0299	.0308	.0527	.0011	.0002
#4	.0278	.0327	.0469	.0000	.0002
#5	.0281	.0301	.0492	.0001	-.0011

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0466	.0270	.0402	.0111
Stddev	.0009	.0011	.0021	.0006
%RSD	1.824	4.014	5.292	5.368
#1	.0470	.0266	.0409	.0116
#2	.0471	.0282	.0431	.0104
#3	.0476	.0274	.0398	.0112
#4	.0455	.0253	.0397	.0106
#5	.0459	.0275	.0373	.0118

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.1527
Stddev	.1993
%RSD	2.4446

#1	8.3226
#2	8.0606
#3	7.8687
#4	8.3549
#5	8.1569

Method: TC_SN_CS Sample Name: ccv

Operator: bjg

Comment: 7.7 ppb Sn

Run Time: 01/25/99 11:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0064	-.0002	6.434	.0002	.0005
Stddev	.0005	.0002	.215	.0002	.0012
%RSD	7.306	109.3	3.335	93.70	240.2

#1	.0060	.0000	6.357	.0000	-.0010
#2	.0062	-.0002	6.350	.0000	.0024
#3	.0063	-.0003	6.525	.0004	.0004
#4	.0072	-.0006	6.185	.0003	.0004
#5	.0063	-.0001	6.752	.0005	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	-.0007	-.0009	-.0029	-.0020
Stddev	.0008	.0006	.0004	.0001
%RSD	107.3	61.50	14.72	4.645

#1	.0005	.0000	-.0030	-.0021
#2	-.0011	-.0016	-.0033	-.0019
#3	-.0003	-.0009	-.0023	-.0020
#4	-.0014	-.0013	-.0028	-.0021
#5	-.0013	-.0009	-.0033	-.0020

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.3803
Stddev	.1572
%RSD	1.8756

#1	8.6080
#2	8.4515
#3	8.3232
#4	8.3277
#5	8.1909

Method: TC_SN_CS Sample Name: ccv

Comment: msprv. lcs

Operator: bjg

Run Time: 01/25/99 11:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	17.27	.0087	1.488	.0000	-.0009
Stddev	.69	.0014	.251	.0005	.0011
%RSD	4.013	16.25	16.87	1186.	129.2
#1	17.94	.0105	1.810	.0007	.0011
#2	17.04	.0078	1.221	.0000	-.0012
#3	17.01	.0095	1.689	-.0005	-.0012
#4	17.98	.0087	1.403	.0003	-.0018
#5	16.36	.0069	1.317	-.0003	-.0013

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	21.89	.0460	.0785	.0529
Stddev	.91	.0033	.0053	.0019
%RSD	4.169	7.216	6.774	3.510
#1	23.28	.0509	.0827	.0542
#2	20.87	.0477	.0718	.0515
#3	22.22	.0439	.0804	.0528
#4	21.42	.0426	.0837	.0552
#5	21.69	.0447	.0740	.0508

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.3774
Stddev	.3305
%RSD	3.9447
#1	8.0981
#2	8.5533
#3	8.4702
#4	7.9831
#5	8.7821

Method: TC_SN_CS Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 01/25/99 11:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[100]	Sn[117]	Sn[126]	Te[125]
Line	99/pulse	100/pulse	117/pulse	126/pulse	125/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0014	.0000	.0012	.0002	-.0001
Stddev	.0012	.0003	.0004	.0006	.0011
%RSD	87.63	2489.	32.92	239.6	778.1
#1	.0034	-.0002	.0018	.0006	-.0002
#2	.0002	.0001	.0012	-.0002	-.0008
#3	.0015	-.0001	.0010	.0006	.0016
#4	.0010	-.0003	.0008	-.0005	-.0011
#5	.0008	.0005	.0015	.0007	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0009	.0003	.0009	.0006
Stddev	.0005	.0007	.0010	.0003
%RSD	49.82	268.4	107.7	49.81
#1	.0012	.0002	-.0005	.0009
#2	.0015	.0004	.0019	.0009
#3	.0010	.0013	.0019	.0007
#4	.0006	-.0001	.0008	.0002
#5	.0004	-.0006	.0005	.0005

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.6455
Stddev	.1904
%RSD	2.2026

#1	8.5492
#2	8.6839
#3	8.4620
#4	8.9553
#5	8.5769

LABCORE Completed Worklist Report for Worklist# 27699

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-5

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000142	1.46e-04	102.817 % Recovery	
1	ICV	0	QMSU-QC U238	QC	0.020	2.06e-02	103.000 % Recovery	
2	ICB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	STD-PREP	0	QMSU-A1 U238-A1	SOLID	10	1.04e+01	104.000 % Recovery	
4	BLNK-PREP	0	QMSU-A1 U233-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U234-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U235-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U236-A1	SOLID	1	<1.60e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U238-A1	SOLID	1	<1.20e-4	ug/g	
5	SAMPLE	S98T003394	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.741e-01	0.474 ug/g	
5	SAMPLE	S98T003394	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.741e-01	0.474 ug/g	
5	SAMPLE	S98T003394	0 A QMSU-A1 U235-A1	SOLID	N/A	< 4.741e-01	0.474 ug/g	
5	SAMPLE	S98T003394	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.321e-01	0.632 ug/g	
5	SAMPLE	S98T003394	0 A QMSU-A1 U238-A1	SOLID	N/A	2.085e+01	0.474 ug/g	
6	DUP	S98T003394	0 A QMSU-A1 U233-A1	SOLID	<4.74e-1	<4.96e-1	RPD	
6	DUP	S98T003394	0 A QMSU-A1 U234-A1	SOLID	<4.74e-1	<4.96e-1	RPD	
6	DUP	S98T003394	0 A QMSU-A1 U235-A1	SOLID	<4.74e-1	<4.96e-1	RPD	
6	DUP	S98T003394	0 A QMSU-A1 U236-A1	SOLID	<6.32e-1	<6.61e-1	RPD	
6	DUP	S98T003394	0 A QMSU-A1 U238-A1	SOLID	2.09e+01	2.57e+01	20.501 RPD	
7	SPK	S98T003394	0 A QMSU-A1 U235-A1	SOLID	14.22	1.53e+01	107.595 % Recovery	
7	SPK	S98T003394	0 A QMSU-A1 U238-A1	SOLID	2000	2.15e+03	107.500 % Recovery	
8	SPK-POST	S98T003394	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.50e-04	105.485 % Recovery	
8	SPK-POST	S98T003394	0 A QMSU-A1 U238-A1	SOLID	0.020	2.09e-02	104.500 % Recovery	
9	SAMPLE	S98T003389	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.945e-01	0.494 ug/g	
9	SAMPLE	S98T003389	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.945e-01	0.494 ug/g	
9	SAMPLE	S98T003389	0 A QMSU-A1 U235-A1	SOLID	N/A	6.098e-01	0.494 ug/g	
9	SAMPLE	S98T003389	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.593e-01	0.659 ug/g	
9	SAMPLE	S98T003389	0 A QMSU-A1 U238-A1	SOLID	N/A	9.209e-01	0.494 ug/g	
10	DUP	S98T003389	0 A QMSU-A1 U233-A1	SOLID	<4.94e-1	<4.58e-1	RPD	
10	DUP	S98T003389	0 A QMSU-A1 U234-A1	SOLID	<4.94e-1	<4.58e-1	RPD	
10	DUP	S98T003389	0 A QMSU-A1 U235-A1	SOLID	6.10e-01	6.37e-01	4.330 RPD	
10	DUP	S98T003389	0 A QMSU-A1 U236-A1	SOLID	<6.59e-1	<6.10e-1	RPD	
10	DUP	S98T003389	0 A QMSU-A1 U238-A1	SOLID	9.21e+01	8.80e+01	4.553 RPD	
11	SPK	S98T003389	0 A QMSU-A1 U235-A1	SOLID	14.22	1.46e+01	102.672 % Recovery	
11	SPK	S98T003389	0 A QMSU-A1 U238-A1	SOLID	2000	2.08e+03	104.000 % Recovery	
12	SPK-POST	S98T003389	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.47e-04	103.376 % Recovery	
12	SPK-POST	S98T003389	0 A QMSU-A1 U238-A1	SOLID	0.020	2.07e-02	103.500 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2611

LABCORE Completed Worklist Report for Worklist# 27699

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
13	CCV	0		QMSU-QC U235	QC	0.000142	1.50e-04	105.634 % Recovery	
13	CCV	0		QMSU-QC U238	QC	0.020	2.17e-02	108.500 % Recovery	
14	CCB	0		QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
14	CCB	0		QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
14	CCB	0		QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
14	CCB	0		QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
14	CCB	0		QMSU-QC U238	QC	1	<1.20e-5	ug/mL	

Final page for worklist# 27699

_____
Analyst Signature   Date

_____
Analyst Signature   Date

3-3-99
Reviewer Signature   Date

12/29/98 08:43

A-0004-1

FILE # U0226A.TFT

Page: 1

LABCORE Data Entry Template for Worklist# 27699

Analyst: BSG Instrument: ICPMS1 Book# 990226-5Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	STD-PREP		@MSU-A1	SOLID		
4	BLANK-PREP		@MSU-A1	SOLID		
5	SAMPLE	S98T003394 0 A	@MSU-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: U233-A1, U234-A1, U235-A1, U236-A1, U238-A1					
6	DUP	S98T003394 0 A	@MSU-A1	SOLID		
7	SPK	S98T003394 0 A	@MSU-A1	SOLID		
8	SPK-POST	S98T003394 0 A	@MSU-A1	SOLID		
9	CCV		@MSU-QC	QC		
10	CCB		@MSU-QC	QC		
~~11~~	~~SAMPLE~~	~~S98T003399 0 A~~	~~@MSU-A1~~	~~SOLID~~	~~98000436~~	~~BY-105 (2)~~
	~~Analytes Requested: U233-A1, U234-A1, U235-A1, U236-A1, U238-A1~~					
~~12~~	~~DUP~~	~~S98T003399 0 A~~	~~@MSU-A1~~	~~SOLID~~		
~~13~~	~~SPK~~	~~S98T003399 0 A~~	~~@MSU-A1~~	~~SOLID~~		
~~14~~	~~SPK-POST~~	~~S98T003399 0 A~~	~~@MSU-A1~~	~~SOLID~~		
15	SAMPLE	S98T003389 0 A	@MSU-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: U233-A1, U234-A1, U235-A1, U236-A1, U238-A1					
16	DUP	S98T003389 0 A	@MSU-A1	SOLID		
17	SPK	S98T003389 0 A	@MSU-A1	SOLID		

Data Entry Comments: RPD for S98T003394 is 20.6% - Sample
will be rerun if requested.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2613

12/29/98 08:43

A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27699

S Type	Sample#	R A	Test	Matrix	Group#	Project
18 SPK-POST	S98T003389	0 A	@MSU-A1	SOLID		
19 CCV			@MSU-QC	QC		
20 CCB			@MSU-QC	QC		

Final page for worklist # 27699

Paul Good 2-26-99
 Analyst Signature Date

Analyst Signature Date

PREP STD .05-10-1
 PREP BLK 1-1-8-1
 S98T003389 .05-10-1

84-D

84-A

84-S

3389 .05-10-1-10-1

89-D

3389

89-A

89-S

3389

3389-D

89-A

89-S

DF

201

10

201

20, 201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

201

Data Entry Comments:

SCANS = 983394, 983395, 983389
 983394, 983395, 983396

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2614

Analysis Report

02/26/99 16:49:24

page 1

Method: MS_TUNE Sample Name: ms-tune Operator: bjg
 Comment:
 Run Time: 02/26/99 16:48 Type: Unk Mode: INT Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	152000.
Stddev	873.
%RSD	.5743

#1	152900.
#2	152900.
#3	152100.
#4	151500.
#5	151300.
#6	153200.
#7	150900.
#8	151200.
#9	151300.
#10	152900.

bjg
 BY-105
 Acid Digest
 ISOU
 File # U02@6A.727
 W.L.# 27699

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2615 TO 2636

Analysis Report

02/26/99 16:59:06

page 1

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 16:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0012	.0015	.1462	.0026	20.56
Stddev	.0010	.0026	.0056	.0008	.70
%RSD	85.33	175.9	3.805	30.35	3.424
#1	-.0005	-.0010	.1395	.0017	19.41
#2	.0019	.0040	.1546	.0027	20.85
#3	.0012	.0046	.1453	.0039	21.29
#4	.0020	-.0005	.1437	.0025	20.75
#5	.0013	.0003	.1477	.0023	20.51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.030
Stddev	.189
%RSD	1.8855

#1	10.326
#2	9.8873
#3	9.8479
#4	10.016
#5	10.073

2616

Method: ISOU Sample Name: ich

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 17:03 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0008	.0003	.0004	.0001
Stddev	.0012	.0011	.0017	.0016	.0014
%RSD	302.6	128.5	579.2	421.5	1887.

#1	-.0005	-.0022	-.0026	.0018	.0007
#2	.0008	.0001	.0011	-.0018	-.0016
#3	-.0024	-.0002	.0015	.0005	-.0012
#4	.0003	-.0001	.0001	.0019	.0013
#5	-.0002	-.0017	.0014	-.0005	.0011

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.176
Stddev	.042
%RSD	.41110

#1	10.199
#2	10.175
#3	10.103
#4	10.204
#5	10.196

Analysis Report

02/26/99 17:08:38

page 1

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 02/26/99 17:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	-.0004	.0005	-.0006	-.0004
Stddev	.0016	.0010	.0015	.0015	.0021
%RSD	399.3	237.4	285.5	234.3	463.0
#1	-.0015	-.0013	.0024	-.0019	.0007
#2	.0006	-.0001	-.0009	-.0019	.0025
#3	.0028	.0011	.0014	-.0009	-.0028
#4	-.0006	-.0006	.0008	-.0015	-.0016
#5	.0007	-.0013	-.0012	-.0007	-.0010

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.252
Stddev	.195
%RSD	1.9040

#1	9.9670
#2	10.346
#3	10.306
#4	10.478
#5	10.161

2618

Analysis Report

02/26/99 17:10:43

page 1

Method: ISOU Sample Name: prepstd

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 17:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0037	.3708	.0069	51.89
Stddev	.0006	.0005	.0057	.0021	.57
%RSD	87.11	13.39	1.539	30.92	1.104

#1	.0001	.0042	.3700	.0065	51.61
#2	-.0015	.0039	.3707	.0063	51.44
#3	-.0004	.0031	.3621	.0043	51.51
#4	-.0008	.0032	.3775	.0102	52.81
#5	-.0007	.0041	.3738	.0074	52.09

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.780
Stddev	.108
%RSD	.99759

#1	10.665
#2	10.765
#3	10.872
#4	10.690
#5	10.908

2619

Analysis Report

02/26/99 17:21:07

page 1

Method: ISOU Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 02/26/99 17:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	-.0022	-.0011	-.0024	.0003
Stddev	.0013	.0012	.0016	.0015	.0003
%RSD	194.5	54.66	142.6	63.50	94.38

#1	-.0004	-.0008	-.0008	-.0001	.0005
#2	-.0010	-.0013	-.0017	-.0015	.0001
#3	-.0005	-.0024	.0005	-.0033	.0003
#4	.0011	-.0025	-.0035	-.0037	-.0001
#5	-.0024	-.0038	-.0001	-.0031	.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.710
Stddev	.097
%RSD	.83160

#1	11.669
#2	11.652
#3	11.827
#4	11.602
#5	11.797

2620

Analysis Report

02/26/99 17:25:43

page 1

Method: ISOU Sample Name: s98t003394 Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 17:24 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0003	.0017	-.0018	.5279
Stddev	.0008	.0005	.0007	.0013	.0105
%RSD	98.62	149.0	44.44	74.25	1.980

#1	-.0011	-.0003	.0010	-.0036	.5195
#2	-.0015	.0004	.0012	-.0012	.5256
#3	-.0016	-.0009	.0017	-.0008	.5212
#4	.0005	-.0008	.0016	-.0006	.5457
#5	-.0006	-.0003	.0029	-.0026	.5276

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.792
Stddev	.142
%RSD	1.3194

#1	10.680
#2	10.911
#3	10.601
#4	10.905
#5	10.863

2621

Analysis Report

02/26/99 17:28:28

page 1

Method: ISOU Sample Name: s98t003394_d Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 17:27 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0018	.0028	-.0015	.6215
Stddev	.0025	.0017	.0008	.0008	.0326
%RSD	938.4	94.57	27.39	51.48	5.247

#1	-.0013	-.0017	.0019	-.0022	.5701
#2	-.0026	-.0011	.0020	-.0018	.6213
#3	-.0017	.0002	.0031	-.0023	.6460
#4	.0038	-.0043	.0033	-.0006	.6172
#5	.0005	-.0018	.0035	-.0008	.6530

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.236
Stddev	.283
%RSD	2.5219

#1	11.692
#2	11.157
#3	10.935
#4	11.281
#5	11.117

2622

Analysis Report

02/26/99 17:30:24

page 1

Method: ISOU Sample Name: s98t003394_a Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 17:29 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0014	.1499	.0008	21.42
Stddev	.0007	.0016	.0056	.0012	.36
%RSD	196.5	117.9	3.769	151.4	1.663

#1	.0007	-.0002	.1475	-.0010	21.35
#2	-.0008	.0008	.1489	.0023	21.86
#3	-.0011	-.0025	.1545	.0004	21.67
#4	-.0002	-.0019	.1422	.0014	21.30
#5	-.0005	-.0031	.1562	.0009	20.94

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.158
Stddev	.212
%RSD	1.9010

#1	11.138
#2	11.031
#3	10.956
#4	11.153
#5	11.509

2623

Analysis Report

02/26/99 17:32:30

page 1

Method: ISOU Sample Name: s98t003394_s Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 17:31 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0010	.0027	.3720	.0043	52.82
Stddev	.0017	.0019	.0067	.0014	1.28
%RSD	174.3	68.65	1.796	32.87	2.426

#1	.0020	.0035	.3722	.0040	53.46
#2	-.0019	.0054	.3706	.0030	53.26
#3	-.0017	.0014	.3638	.0034	51.67
#4	-.0013	.0026	.3710	.0045	51.34
#5	-.0018	.0007	.3824	.0066	54.39

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.350
Stddev	.167
%RSD	1.4673

#1	11.281
#2	11.256
#3	11.488
#4	11.560
#5	11.165

2624

Analysis Report

02/26/99 17:38:24

page 1

Method: ISOU Sample Name: ccv Operator: bjg
 Comment: iso_u lcs
 Run Time: 02/26/99 17:37 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0011	.0005	.1538	.0005	21.67
Stddev	.0017	.0007	.0105	.0016	.59
%RSD	160.2	130.1	6.808	313.8	2.715

#1	-.0038	.0006	.1433	-.0002	20.82
#2	-.0005	.0002	.1644	-.0003	22.14
#3	.0008	-.0001	.1615	.0033	22.09
#4	-.0013	.0017	.1420	-.0006	21.30
#5	-.0005	.0003	.1580	.0003	22.02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.807
Stddev	.222
%RSD	2.0541

#1	11.190
#2	10.650
#3	10.679
#4	10.805
#5	10.711

2625

Analysis Report

02/26/99 17:44:26

page 1

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 17:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0010	-.0013	-.0014	-.0008	-.0006
Stddev	.0022	.0011	.0013	.0010	.0015
%RSD	232.1	84.76	93.27	135.2	271.0

#1	-.0006	-.0011	-.0019	.0005	-.0005
#2	-.0014	-.0001	-.0019	-.0017	-.0014
#3	.0004	-.0028	-.0028	.0001	.0002
#4	-.0045	-.0020	.0008	-.0014	.0014
#5	.0014	-.0005	-.0013	-.0014	-.0026

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.703
Stddev	.080
%RSD	.74383

#1	10.568
#2	10.719
#3	10.771
#4	10.752
#5	10.707

2626

Analysis Report

02/26/99 18:08:22

page 1

Method: ISOU   Sample Name: s98t003399   Operator: bjg
 Comment: 20301 df
 Run Time: 02/26/99 18:07 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0017	-.0011	.0669	-.0011	10.34
Stddev	.0011	.0017	.0027	.0006	.14
%RSD	66.45	159.3	4.001	58.52	1.375

#1	-.0021	-.0025	.0640	-.0011	10.21
#2	-.0018	.0013	.0711	-.0004	10.21
#3	-.0034	-.0020	.0653	-.0019	10.45
#4	-.0006	.0002	.0670	-.0006	10.31
#5	-.0006	-.0025	.0671	-.0014	10.52

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.809
Stddev	.280
%RSD	2.5899

#1	11.245
#2	10.689
#3	10.524
#4	10.909
#5	10.677

2627

Analysis Report

02/26/99 18:14:33

page 1

Method: ISOU   Sample Name: s98t003399_d   Operator: bjg
 Comment: 20301 df
 Run Time: 02/26/99 18:13 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0003	.1531	.0028	23.39
Stddev	.0012	.0012	.0056	.0017	.44
%RSD	3044.	377.3	3.670	60.13	1.868

#1	-.0001	.0024	.1453	.0010	23.04
#2	.0008	-.0008	.1517	.0029	24.12
#3	.0002	.0000	.1579	.0019	23.17
#4	.0013	-.0003	.1592	.0055	23.47
#5	-.0020	.0003	.1513	.0027	23.15

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.643
Stddev	.103
%RSD	.96670

#1	10.800
#2	10.546
#3	10.649
#4	10.555
#5	10.664

2628

Analysis Report

02/26/99 18:17:34

page 1

Method: ISOU Sample Name: s98t003399_a Operator: bjg
 Comment: 20301 df
 Run Time: 02/26/99 18:16 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0017	-.0016	.1132	.0014	17.63
Stddev	.0014	.0004	.0041	.0012	.28
%RSD	84.15	25.19	3.612	85.82	1.593

#1	-.0008	-.0022	.1160	.0018	17.54
#2	-.0037	-.0017	.1107	-.0007	17.41
#3	-.0004	-.0014	.1165	.0019	17.45
#4	-.0026	-.0014	.1072	.0021	17.66
#5	-.0010	-.0012	.1156	.0021	18.10

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.613
Stddev	.146
%RSD	1.3773

#1	10.695
#2	10.675
#3	10.462
#4	10.778
#5	10.454

2629

Analysis Report

02/26/99 18:20:24

page 1

Method: ISOU Sample Name: s98t003399_s Operator: bjg
 Comment: 20301 df
 Run Time: 02/26/99 18:19 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0003	.0807	.0007	12.26
Stddev	.0007	.0007	.0034	.0016	.32
%RSD	81.09	233.9	4.241	225.7	2.588

#1	-.0015	-.0004	.0753	.0019	11.91
#2	-.0003	.0008	.0848	.0007	12.43
#3	-.0015	-.0005	.0808	-.0020	11.94
#4	.0000	-.0003	.0814	.0015	12.38
#5	-.0010	-.0010	.0813	.0014	12.63

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.452
Stddev	.106
%RSD	1.0102

#1	10.543
#2	10.339
#3	10.473
#4	10.560
#5	10.345

2630

Analysis Report

02/26/99 18:25:47

page 1

Method: ISOU Sample Name: s98t003389 Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 18:24 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0015	-.0023	.0148	-.0018	2.235
Stddev	.0015	.0013	.0005	.0013	.068
%RSD	97.96	57.90	3.514	72.21	3.060

#1	.0014	-.0015	.0142	-.0029	2.149
#2	.0030	-.0023	.0155	-.0015	2.227
#3	-.0008	-.0006	.0146	-.0007	2.194
#4	.0024	-.0041	.0146	-.0005	2.319
#5	.0014	-.0028	.0150	-.0034	2.284

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.741
Stddev	.149
%RSD	1.3846

#1	10.808
#2	10.681
#3	10.959
#4	10.569
#5	10.685

2631

Analysis Report

02/26/99 18:28:28

page 1

Method: ISOU Sample Name: s98t003389_d Operator: bjg
 Comment: 201 df
 Run Time: 02/28/99 18:27 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	-.0025	.0167	-.0012	2.307
Stddev	.0016	.0008	.0027	.0013	.060
%RSD	1099.	31.34	16.41	106.2	2.586
#1	.0003	-.0024	.0167	-.0028	2.362
#2	-.0018	-.0037	.0159	.0003	2.340
#3	.0018	-.0025	.0215	-.0002	2.233
#4	.0007	-.0016	.0147	-.0014	2.349
#5	-.0018	-.0020	.0150	-.0019	2.253

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.375
Stddev	.165
%RSD	1.4533

#1	11.203
#2	11.291
#3	11.477
#4	11.292
#5	11.610

2632

Analysis Report

02/26/99 18:31:20

page 1

Method: ISOU

Sample Name: s98t003389_a

Operator: bjg

Comment: 201 df

Run Time: 02/26/99 18:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	-.0007	.1619	.0004	22.95
Stddev	.0014	.0006	.0029	.0012	.62
%RSD	352.8	90.28	1.820	298.9	2.713
#1	-.0017	-.0010	.1623	.0019	23.14
#2	.0010	-.0011	.1582	-.0007	21.85
#3	-.0003	-.0008	.1658	-.0003	23.24
#4	.0018	.0004	.1633	.0016	23.10
#5	.0012	-.0010	.1600	-.0004	23.40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.501
Stddev	.057
%RSD	.49873

#1	11.493
#2	11.598
#3	11.467
#4	11.498
#5	11.451

2633

Analysis Report

02/26/99 18:33:24

page 1

Method: ISOU Sample Name: s98t003389_s Operator: bjg
 Comment: 201 df
 Run Time: 02/26/99 18:32 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0021	.3908	.0054	55.93
Stddev	.0016	.0028	.0257	.0024	4.00
%RSD	522.7	138.5	6.580	44.11	7.145

#1	-.0016	-.0014	.3697	.0034	52.56
#2	-.0001	.0005	.3810	.0031	54.42
#3	.0003	.0017	.3883	.0067	55.88
#4	.0029	.0035	.3799	.0048	54.02
#5	.0000	.0060	.4353	.0088	62.76

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.814
Stddev	.645
%RSD	5.4553

#1	12.247
#2	12.086
#3	11.832
#4	12.208
#5	10.698

2634

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 18:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0007	-.0011	.1497	-.0002	21.72
Stddev	.0014	.0008	.0039	.0020	.67
%RSD	192.6	71.42	2.629	1153.	3.107
#1	-.0005	-.0022	.1535	-.0004	21.58
#2	-.0011	-.0012	.1433	-.0013	20.96
#3	.0014	-.0007	.1523	.0019	22.78
#4	-.0024	-.0013	.1501	-.0029	21.83
#5	-.0009	-.0001	.1494	.0017	21.45

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.486
Stddev	.267
%RSD	2.3214

#1	11.549
#2	11.771
#3	11.059
#4	11.440
#5	11.610

Analysis Report

Method: ISOU

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 18:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0022	-.0025	-.0013	-.0023	-.0022
Stddev	.0014	.0009	.0009	.0011	.0011
%RSD	65.12	35.28	65.80	48.68	48.94
#1	-.0034	-.0030	-.0022	-.0033	-.0030
#2	-.0019	-.0027	-.0007	-.0033	-.0032
#3	-.0037	-.0036	-.0022	-.0006	-.0028
#4	-.0019	-.0022	-.0003	-.0017	-.0011
#5	-.0001	-.0012	-.0012	-.0026	-.0010

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.303
Stddev	.193
%RSD	1.7099
#1	11.520
#2	11.328
#3	11.036
#4	11.438
#5	11.192

LABCORE Completed Worklist Report for Worklist# 27701

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-5

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSU-QC U235	QC	0.000142	1.42e-04	100.000 % Recovery	
1	ICV	0	EMSU-QC U238	QC	0.020	2.08e-02	104.000 % Recovery	
2	ICB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	SAMPLE	S98T003385 0	D EMSU-D1 U233-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T003385 0	D EMSU-D1 U234-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T003385 0	D EMSU-D1 U235-D1	LIQUID	N/A	< 4.860e-01	0.486 ug/mL	
3	SAMPLE	S98T003385 0	D EMSU-D1 U236-D1	LIQUID	N/A	< 6.480e-01	0.648 ug/mL	
3	SAMPLE	S98T003385 0	D EMSU-D1 U238-D1	LIQUID	N/A	1.284e+00	0.486 ug/mL	
4	DUP	S98T003385 0	D EMSU-D1 U233-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T003385 0	D EMSU-D1 U234-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T003385 0	D EMSU-D1 U235-D1	LIQUID	<4.86e-1	<4.86e-1	RPD	
4	DUP	S98T003385 0	D EMSU-D1 U236-D1	LIQUID	<6.48e-1	<6.48e-1	RPD	
4	DUP	S98T003385 0	D EMSU-D1 U238-D1	LIQUID	1.28e+00	1.16e+00	9.836 RPD	
5	SPK-POST	S98T003385 0	D EMSU-D1 U235-D1	LIQUID	0.0001422	1.35e-04	94.937 % Recovery	
5	SPK-POST	S98T003385 0	D EMSU-D1 U238-D1	LIQUID	0.020	1.98e-02	99.000 % Recovery	
6	CCV	0	EMSU-QC U235	QC	0.000142	1.46e-04	102.817 % Recovery	
6	CCV	0	EMSU-QC U238	QC	0.020	2.08e-02	104.000 % Recovery	
7	CCB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
7	CCB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
7	CCB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
7	CCB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
7	CCB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	

Final page for worklist# 27701

Analyst Signature

Date

Analyst Signature

Date

Saul M. Pang 03/01/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27701

Analyst: BIG Instrument: ICPMS1 Book# 220226-5

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 SAMPLE	S98T003385	0 D	@MSU-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
4 DUP	S98T003385	0 D	@MSU-D1	LIQUID		
5 SPK-POST	S98T003385	0 D	@MSU-D1	LIQUID		
6 CCV			@MSU-QC	QC		
7 CCB			@MSU-QC	QC		

Final page for worklist # 27701

Saul Green 276-88
Analyst Signature Date

Analyst Signature Date

S98T003385 .1-10-.025-10 DF 80,501
3385-D
3385-A ↓ ↓

Top post spike = .02 of 10 pp USTD in 10 ml

Data Entry Comments:

SCAN = 883385

Upload & Validated by:

Saul M. Pang
03/01/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2638

Analysis Report

02/26/99 20:03:27

page 1

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 02/26/99

20:02 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	170400.
Stddev	1158.
%RSD	.6794

#1	170200.
#2	171100.
#3	172400.
#4	171600.
#5	168900.
#6	169900.
#7	171400.
#8	169900.
#9	169000.
#10	169800.

Effend
BY-105

5887003385

.1-10, 025-10

FILE # 002265.777

W.L.# 27701

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT SKW 3/26/99
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2639 TO 26543

Analysis Report

02/26/99 20:05:42

page 1

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/26/99 20:04 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.980	5.709	6.146	6.243	6.036
Stddev	.373	.884	.308	.705	.584
%RSD	6.240	15.49	5.009	11.28	9.682

#1	5.626	4.439	6.021	5.670	6.461
#2	6.422	6.378	6.160	7.470	5.898
#3	6.065	5.502	6.585	5.978	5.155
#4	6.222	6.704	5.740	6.090	6.660
#5	5.567	5.523	6.225	6.005	6.005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.436
Stddev	.064
%RSD	.55997

#1	11.376
#2	11.445
#3	11.542
#4	11.411
#5	11.406

2640

Method : ISOU

Standardzn Report

02/26/99 20:05:46

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	457.1511	5.9803	1.0000000	02/26/99 20:05:39

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.9803
100ppb U	0.711	0.711	0	0	331.01

U [233]
0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	457.5324	5.7092	1.0000000	02/26/99 20:05:39

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.7092
100ppb U	0.711	0.711	0	0	331.01

U [234]
0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	456.9180	6.1461	1.0000000	02/26/99 20:05:39

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.1461
100ppb U	0.711	0.711	0	0	331.01

U [235]
0
(S) IR³⁷⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	456.7819	6.2428	1.0000000	02/26/99 20:05:39

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.2428
100ppb U	0.711	0.711	0	0	331.01

U [236]
0
(S) IR³⁷⁰⁰¹⁰

Concentration

2641

Method : ISOU

Standardzn Report

02/26/99 20:05:46

page 2

El Name	Slope	Y-int	Correlation	Date Stdized
U [238]	413.8376	6.0358	1.0000000	02/26/99 20:05:39

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	6.0358
100ppb U	99.3	99.3	0	0	41100

U [238]

0

(S) IR

Concentration

Method: ISOU Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/26/99 20:08 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	343.3	343.3	343.3	343.3	42880.
Stddev	6.8	6.8	6.8	6.8	929.
%RSD	1.972	1.972	1.972	1.972	2.167

#1	337.9	337.9	337.9	337.9	42090.
#2	348.6	348.6	348.6	348.6	44110.
#3	336.6	336.6	336.6	336.6	41920.
#4	352.1	352.1	352.1	352.1	43500.
#5	341.3	341.3	341.3	341.3	42760.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	11.209
Stddev	.177
%RSD	1.5812

#1	11.310
#2	10.938
#3	11.311
#4	11.121
#5	11.363

El Name	Slope	Y-int	Correlation	Date Stdized
U_ ₀ [233]	474.4170	5.9803	1.0000000	02/26/99 20:09:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.9803
100ppb U	0.711	0.711	0	0	343.29

U₀ [233]
(S) IR³⁸⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ ₀ [234]	474.7983	5.7092	1.0000000	02/26/99 20:09:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	5.7092
100ppb U	0.711	0.711	0	0	343.29

U₀ [234]
(S) IR³⁸⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ ₀ [235]	474.1839	6.1461	1.0000000	02/26/99 20:09:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.1461
100ppb U	0.711	0.711	0	0	343.29

U₀ [235]
(S) IR³⁸⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ ₀ [236]	474.0478	6.2428	1.0000000	02/26/99 20:09:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.2428
100ppb U	0.711	0.711	0	0	343.29

U₀ [236]
(S) IR³⁸⁰⁰¹⁰

Concentration

Method : ISOU

Standardzn Report

02/26/99 20:09:28

page 2

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	431.7166	6.0358	1.0000000	02/26/99 20:09:20

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	6.0358
100ppb U	99.3	99.3	0	0	42875

U [238]

0

47170110
(S) IR

Concentration

2645

Analysis Report

02/26/99 20:13:29

page 1

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 02/26/99 20:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0010	-.0006	-.0011	-.0009	.0001
Stddev	.0014	.0011	.0006	.0012	.0010
%RSD	139.2	168.7	51.87	134.6	913.1

#1	.0003	-.0008	-.0006	.0008	.0010
#2	-.0023	.0006	-.0006	-.0023	.0000
#3	-.0013	-.0003	-.0009	-.0016	.0010
#4	-.0024	-.0023	-.0018	-.0002	-.0001
#5	.0007	-.0004	-.0016	-.0012	-.0014

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	11.787
Stddev	.180
%RSD	1.5257

#1	11.716
#2	11.503
#3	11.879
#4	11.931
#5	11.907

2646

Analysis Report

02/26/99 20:16:27

page 1

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 20:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	.0011	.1420	.0004	20.75
Stddev	.0011	.0013	.0040	.0029	.22
%RSD	165.3	119.0	2.790	686.4	1.078

#1	.0008	.0023	.1417	.0037	20.77
#2	-.0004	.0009	.1465	.0021	21.05
#3	-.0021	-.0010	.1387	-.0039	20.42
#4	-.0011	.0013	.1456	-.0003	20.77
#5	-.0004	.0020	.1377	.0006	20.75

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	11.672
Stddev	.076
%RSD	.65120

#1	11.581
#2	11.628
#3	11.784
#4	11.694
#5	11.675

2647

Analysis Report

02/26/99 20:22:19

page 1

Method: ISOU Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 20:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0011	-.0007	-.0011	-.0016	-.0005
Stddev	.0006	.0013	.0007	.0008	.0009
%RSD	53.05	176.6	66.72	51.14	199.0

#1	-.0008	-.0013	-.0016	-.0007	-.0006
#2	-.0012	.0009	-.0021	-.0028	-.0012
#3	-.0007	-.0004	-.0005	-.0018	-.0007
#4	-.0022	-.0027	-.0005	-.0010	-.0009
#5	-.0009	-.0003	-.0007	-.0014	.0011

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	11.980
Stddev	.261
%RSD	2.1789

#1	11.535
#2	12.163
#3	12.104
#4	12.139
#5	11.958

2648

Analysis Report

02/26/99 20:29:04

page 1

Method: ISOU Sample Name: s98t003385

Operator: bjg

Comment: 40501 df

Run Time: 02/26/99 20:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0014	-.0014	-.0011	-.0022	.0317
Stddev	.0008	.0004	.0009	.0008	.0017
%RSD	57.24	31.39	76.49	37.52	5.340

#1	-.0001	-.0012	-.0013	-.0020	.0339
#2	-.0013	-.0018	-.0012	-.0016	.0297
#3	-.0016	-.0008	-.0014	-.0024	.0326
#4	-.0019	-.0012	-.0021	-.0036	.0319
#5	-.0020	-.0017	.0003	-.0016	.0303

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.094
Stddev	.137
%RSD	1.1303

#1	12.173
#2	11.880
#3	12.207
#4	12.034
#5	12.175

2649

Analysis Report

02/26/99 20:31:55

page 1

Method: ISOU   Sample Name: s98t003385_d   Operator: bjg
 Comment: 40501 df
 Run Time: 02/26/99 20:30 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0013	-.0010	-.0018	-.0018	.0286
Stddev	.0009	.0010	.0005	.0007	.0022
%RSD	67.54	102.5	28.09	38.84	7.526
#1	.0000	.0001	-.0017	-.0011	.0307
#2	-.0020	-.0001	-.0019	-.0020	.0275
#3	-.0008	-.0017	-.0012	-.0021	.0292
#4	-.0016	-.0009	-.0026	-.0010	.0255
#5	-.0019	-.0022	-.0019	-.0026	.0302
Int. Std.	Ir[193]				
Line	193/pulse				
Units	Cts/s				
Avg	12.391				
Stddev	.127				
%RSD	1.0213				
#1	12.276				
#2	12.246				
#3	12.438				
#4	12.449				
#5	12.547				

2650

Analysis Report

02/26/99 20:34:23

page 1

Method: ISOU   Sample Name: s98t003385_a   Operator: bjg
 Comment: 40501 df
 Run Time: 02/26/99 20:33 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	- .0012	.0011	.1346	.0005	19.85
Stddev	.0009	.0006	.0054	.0008	.80
%RSD	68.96	55.40	4.003	165.2	4.008

#1	- .0019	.0013	.1356	- .0005	19.36
#2	- .0012	.0015	.1261	- .0001	19.29
#3	- .0021	.0005	.1333	.0014	19.18
#4	- .0012	.0019	.1382	.0010	20.65
#5	.0001	.0004	.1399	.0005	20.79

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.710
Stddev	.344
%RSD	2.7045

#1	12.687
#2	12.892
#3	13.188
#4	12.447
#5	12.334

2651

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso_u lcs

Run Time: 02/26/99 20:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0018	-.0001	.1455	-.0003	20.79
Stddev	.0006	.0006	.0048	.0016	.44
%RSD	34.07	531.1	3.293	537.7	2.109
#1	-.0015	-.0002	.1519	-.0018	21.17
#2	-.0024	-.0004	.1467	.0018	21.05
#3	-.0009	.0009	.1473	.0009	21.09
#4	-.0020	-.0007	.1400	-.0014	20.41
#5	-.0020	-.0002	.1414	-.0010	20.23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.966
Stddev	.254
%RSD	1.9553

#1	12.634
#2	12.861
#3	12.926
#4	13.101
#5	13.307

Analysis Report

02/26/99 20:41:50

page 1

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 02/26/99 20:40 Type: Unk

Mode: CONC

Corr. Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0027	-.0009	-.0027	-.0027	-.0012
Stddev	.0004	.0005	.0008	.0012	.0017
%RSD	13.66	57.42	29.48	46.03	141.1

#1	-.0029	-.0012	-.0025	-.0031	.0003
#2	-.0023	-.0001	-.0039	-.0031	-.0026
#3	-.0024	-.0013	-.0026	-.0043	-.0002
#4	-.0029	-.0013	-.0028	-.0009	-.0001
#5	-.0032	-.0007	-.0017	-.0022	-.0034

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.528
Stddev	.267
%RSD	1.9705

#1	13.105
#2	13.445
#3	13.614
#4	13.773
#5	13.704

LABCORE Completed Worklist Report for Worklist# 27705

Analyst: kjt

Instrument: ICPMS1

Book#: 990223-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSU-QC U235	QC	0.000142	1.40e-04	98.592 % Recovery	
1	ICV	0	EMSU-QC U238	QC	0.020	2.02e-02	101.000 % Recovery	
2	ICB	0	EMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	EMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	EMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	STD-PREP	0	EMSU-A1 U238-A1	SOLID	10.0	9.45e0	94.500 % Recovery	
4	BLNK-PREP	0	EMSU-A1 U233-A1	SOLID	1	<1.32e-4	ug/g	
4	BLNK-PREP	0	EMSU-A1 U234-A1	SOLID	1	<1.32e-4	ug/g	
4	BLNK-PREP	0	EMSU-A1 U235-A1	SOLID	1	<1.32e-4	ug/g	
4	BLNK-PREP	0	EMSU-A1 U236-A1	SOLID	1	<1.76e-4	ug/g	
4	BLNK-PREP	0	EMSU-A1 U238-A1	SOLID	1	<1.32e-4	ug/g	
5	SAMPLE	S98T002713	0 A EMSU-A1 U233-A1	SOLID	N/A	< 4.786e-01	0.479 ug/g	
5	SAMPLE	S98T002713	0 A EMSU-A1 U234-A1	SOLID	N/A	< 4.786e-01	0.479 ug/g	
5	SAMPLE	S98T002713	0 A EMSU-A1 U235-A1	SOLID	N/A	2.467e+01	0.479 ug/g	
5	SAMPLE	S98T002713	0 A EMSU-A1 U236-A1	SOLID	N/A	6.740e-01	0.638 ug/g	
5	SAMPLE	S98T002713	0 A EMSU-A1 U238-A1	SOLID	N/A	3.578e+03	0.479 ug/g	
6	DUP	S98T002713	0 A EMSU-A1 U233-A1	SOLID	<4.79e-1	<4.69e-1	RPD	
6	DUP	S98T002713	0 A EMSU-A1 U234-A1	SOLID	<4.79e-1	<4.69e-1	RPD	
6	DUP	S98T002713	0 A EMSU-A1 U235-A1	SOLID	2.47e+01	3.08e+01	21.982 RPD	
6	DUP	S98T002713	0 A EMSU-A1 U236-A1	SOLID	6.74e-01	8.41e-01	22.046 RPD	
6	DUP	S98T002713	0 A EMSU-A1 U238-A1	SOLID	3.58e+03	4.41e+03	20.776 RPD	
7	SAMPLE	S98T002846	0 A EMSU-A1 U233-A1	SOLID	N/A	< 4.918e-01	0.492 ug/g	
7	SAMPLE	S98T002846	0 A EMSU-A1 U234-A1	SOLID	N/A	< 4.918e-01	0.492 ug/g	
7	SAMPLE	S98T002846	0 A EMSU-A1 U235-A1	SOLID	N/A	1.500e+01	0.492 ug/g	
7	SAMPLE	S98T002846	0 A EMSU-A1 U236-A1	SOLID	N/A	< 6.558e-01	0.656 ug/g	
7	SAMPLE	S98T002846	0 A EMSU-A1 U238-A1	SOLID	N/A	2.299e+03	0.492 ug/g	
8	DUP	S98T002846	0 A EMSU-A1 U233-A1	SOLID	<4.92e-1	<4.87e-1	RPD	
8	DUP	S98T002846	0 A EMSU-A1 U234-A1	SOLID	<4.92e-1	<4.87e-1	RPD	
8	DUP	S98T002846	0 A EMSU-A1 U235-A1	SOLID	1.50e+01	1.54e+01	2.632 RPD	
8	DUP	S98T002846	0 A EMSU-A1 U236-A1	SOLID	<6.56e-1	<6.49e-1	RPD	
8	DUP	S98T002846	0 A EMSU-A1 U238-A1	SOLID	2.30e+03	2.34e+03	1.724 RPD	
9	SPK	S98T002846	0 A EMSU-A1 U235-A1	SOLID	14.22	1.39e+01	97.750 % Recovery	
9	SPK	S98T002846	0 A EMSU-A1 U238-A1	SOLID	2000.00	1.81e+03	90.500 % Recovery	
10	SPK-POST	S98T002846	0 A EMSU-A1 U235-A1	SOLID	0.0001422	1.39e-04	97.750 % Recovery	
10	SPK-POST	S98T002846	0 A EMSU-A1 U238-A1	SOLID	0.020	1.74e-02	87.000 % Recovery	
11	SAMPLE	S98T002740	0 A EMSU-A1 U233-A1	SOLID	N/A	< 4.924e-01	0.492 ug/g	
11	SAMPLE	S98T002740	0 A EMSU-A1 U234-A1	SOLID	N/A	< 4.924e-01	0.492 ug/g	
11	SAMPLE	S98T002740	0 A EMSU-A1 U235-A1	SOLID	N/A	1.330e+00	0.492 ug/g	
11	SAMPLE	S98T002740	0 A EMSU-A1 U236-A1	SOLID	N/A	< 6.566e-01	0.657 ug/g	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27705

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11 SAMPLE	S98T002740	0 A	QMSU-A1 U238-A1	SOLID	N/A	1.875e+02	0.492	ug/g
12 DUP	S98T002740	0 A	QMSU-A1 U233-A1	SOLID	<4.92e-1	<4.68e-1		RPD
12 DUP	S98T002740	0 A	QMSU-A1 U234-A1	SOLID	<4.92e-1	<4.68e-1		RPD
12 DUP	S98T002740	0 A	QMSU-A1 U235-A1	SOLID	1.33e+00	1.51e+00	12.676	RPD
12 DUP	S98T002740	0 A	QMSU-A1 U236-A1	SOLID	<6.57e-1	<6.24e-1		RPD
12 DUP	S98T002740	0 A	QMSU-A1 U238-A1	SOLID	1.88e+02	2.18e+02	14.778	RPD
13 SPK	S98T002740	0 A	QMSU-A1 U235-A1	SOLID	14.22	1.37e+01	96.343	% Recovery
13 SPK	S98T002740	0 A	QMSU-A1 U238-A1	SOLID	2000.0	1.81e+03	90.500	% Recovery
14 SPK-POST	S98T002740	0 A	QMSU-A1 U235-A1	SOLID	0.0001422	1.29e-04	90.717	% Recovery
14 SPK-POST	S98T002740	0 A	QMSU-A1 U238-A1	SOLID	0.020	1.77e-02	88.500	% Recovery
15 CCV		0	QMSU-QC U235	QC	0.000142	1.38e-04	97.183	% Recovery
15 CCV		0	QMSU-QC U238	QC	0.020	1.93e-02	96.500	% Recovery
16 CCB		0	QMSU-QC U233	QC	1	<1.20e-5		ug/mL
16 CCB		0	QMSU-QC U234	QC	1	<1.20e-5		ug/mL
16 CCB		0	QMSU-QC U235	QC	1	<1.20e-5		ug/mL
16 CCB		0	QMSU-QC U236	QC	1	<1.60e-5		ug/mL
16 CCB		0	QMSU-QC U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 27705

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul H. Pang 2/26/99
_____
Reviewer Signature Date

12/29/98 09:03

A-0004-1

LABCORE Data Entry Template for Worklist# 27705

Analyst: KST Instrument: ICPMS1 Book# 990223-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV				@MSU-QC	QC		
2	ICB				@MSU-QC	QC		
3	STD-PREP				@MSU-A1	SOLID		
4	BLNK-PREP				@MSU-A1	SOLID		
5	SAMPLE	S98T002713	0	A	@MSU-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1								
6	DUP	S98T002713	0	A	@MSU-A1	SOLID		
7	SAMPLE	S98T002846	0	A	@MSU-A1	SOLID	98000437	BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1								
8	DUP	S98T002846	0	A	@MSU-A1	SOLID		
9	SPK	S98T002846	0	A	@MSU-A1	SOLID		
10	SPK-POST	S98T002846	0	A	@MSU-A1	SOLID		
11	~~CCV~~				@MSU-QC	QC		
12	~~CCB~~				@MSU-QC	QC		
Not Needed								
13	SAMPLE	S98T002740	0	A	@MSU-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1								
14	DUP	S98T002740	0	A	@MSU-A1	SOLID		
15	SPK	S98T002740	0	A	@MSU-A1	SOLID		
16	SPK-POST	S98T002740	0	A	@MSU-A1	SOLID		
17	CCV				@MSU-QC	QC		

Data Entry Comments:

Uploaded & Validated by
 Karl A. King
 2/26/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

12/29/98 09:03

Page: 2

A-0004-1

LABCORE Data Entry Template for Worklist# 27705

S Type	Sample#	R A	Test	Matrix	Group#	Project
18 CCB			@MSU-QC	QC		

Final page for worklist # 27705

Ks Thorndike 02-25-99

Analyst Signature

Date

Analyst Signature

Date

File # 0225b.txt

SH-Prep 05-10 201

Blk-Prep 4-10 11

s98+2713 05-10 201

2740-S 05-10 201

2740-a 05-10 201

2713-d

2846

2846-d

2846-S

2846-a

2740

2740-d

Mass Scan 190-250

98+2713

98+2846

98+2740

Used .02 ml ~~8990212-1~~
 100 ppm Uranium for spike
 Ks 02-25-99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2657

Method: MS_TUNE Sample Name: MS_TUNE

Operator:

Comment:

Run Time: 02/25/99 14:29 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/s
Avg	206600.
Stddev	2118.
%RSD	1.025

#1	204700.
#2	206100.
#3	205800.
#4	205300.
#5	203700.
#6	209300.
#7	208100.
#8	204800.
#9	209200.
#10	209000.

SP
2/25/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2658 TO 2682.

Method: MS_TUNE Sample Name: MS_TUNE

Operator:

Comment:

Run Time: 02/25/99 14:30 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	209400.
Stddev	1819.
%RSD	.8688

#1	207400.
#2	207500.
#3	208500.
#4	210300.
#5	210000.
#6	210900.
#7	207200.
#8	208700.
#9	211000.
#10	212500.

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/25/99 14:39 Type: Std

Mode: IR

Corr. Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.490	3.194	3.054	3.484	3.430
Stddev	.204	.566	.291	.219	.376
%RSD	5.850	17.71	9.533	6.286	10.95
#1	3.258	3.661	2.956	3.157	2.788
#2	3.373	2.844	3.472	3.638	3.737
#3	3.431	3.365	2.936	3.398	3.431
#4	3.618	2.390	3.198	3.715	3.618
#5	3.772	3.708	2.708	3.514	3.579

Int. Std.

Ir[193]

Line 193/pulse

Units Cts/S

Avg 15.230

Stddev .263

%RSD 1.7247

#1	14.886
#2	15.119
#3	15.158
#4	15.479
#5	15.509

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	406.5551	3.4903	1.0000000	02/25/99 14:41:11

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	3.4903
100ppb U	0.711	0.711	0	0	292.55

U [233]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	406.9725	3.1935	1.0000000	02/25/99 14:41:11

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	3.1935
100ppb U	0.711	0.711	0	0	292.55

U [234]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	407.1688	3.0540	1.0000000	02/25/99 14:41:11

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	3.054
100ppb U	0.711	0.711	0	0	292.55

U [235]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	406.5636	3.4843	1.0000000	02/25/99 14:41:11

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	3.4843
100ppb U	0.711	0.711	0	0	292.55

U [236]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_{238}	394.4296	3.4304	1.0000000	02/25/99 14:41:11

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S)IR
Blank	0	0	0	0	3.4304
100ppb U	99.3	99.3	0	0	39170

U   [238]
0
43090110
(S)IR   Concentration

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/25/99 14:43 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	307.4	307.4	307.4	307.4	39070.
Stddev	8.2	8.2	8.2	8.2	553.
%RSD	2.679	2.679	2.679	2.679	1.416
#1	293.9	293.9	293.9	293.9	38300.
#2	306.8	306.8	306.8	306.8	39410.
#3	313.1	313.1	313.1	313.1	39760.
#4	308.3	308.3	308.3	308.3	39010.
#5	314.9	314.9	314.9	314.9	38890.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.357
Stddev	.229
%RSD	1.4917

#1	15.766
#2	15.245
#3	15.278
#4	15.235
#5	15.260

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	427.4554	3.4903	1.0000000	02/25/99 14:44:27

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	3.4903
100ppb U	0.711	0.711	0	0	307.41

U [233]

0
(S) IR³⁴⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	427.8728	3.1935	1.0000000	02/25/99 14:44:27

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	3.1935
100ppb U	0.711	0.711	0	0	307.41

U [234]

0
(S) IR³⁴⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	428.0691	3.0540	1.0000000	02/25/99 14:44:27

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	3.054
100ppb U	0.711	0.711	0	0	307.41

U [235]

0
(S) IR³⁴⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	427.4638	3.4843	1.0000000	02/25/99 14:44:27

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	3.4843
100ppb U	0.711	0.711	0	0	307.41

U [236]

0
(S) IR³⁴⁰⁰¹⁰

Concentration

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	393.4395	3.4304	1.0000000	02/25/99 14:44:27

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	3.4304
100ppb U	99.3	99.3	0	0	39072

U [238]

(S) IR

Concentration

Method: ISOU

Sample Name: blank

Operator: kjt

Comment: background

Run Time: 02/25/99 14:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0011	-.0004	.0000	-.0011	.0002
Stddev	.0006	.0006	.001	.0006	.0008
%RSD	55.17	158.7	1350.	55.91	482.8

#1	-.0003	.0004	.0008	-.0003	.0011
#2	-.0017	-.0010	-.0007	-.0017	-.0005
#3	-.0011	-.0004	-.0001	-.0011	.0007
#4	-.0007	.0000	.0003	-.0007	.0004
#5	-.0015	-.0008	-.0005	-.0015	-.0008

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.218
Stddev	.528
%RSD	3.4702

#1	14.659
#2	14.746
#3	15.583
#4	15.887
#5	15.215

Method: ISOU

Sample Name: icv

Operator: kjt

Comment: iso-u lcs

Run Time: 02/25/99 14:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0011	.0000	.1401	.0015	20.22
Stddev	.0008	.0005	.0059	.0010	.39
%RSD	71.45	1170.	4.195	65.24	1.908
#1	-.0006	-.0002	.1498	.0024	20.74
#2	-.0002	-.0004	.1390	.0025	20.19
#3	-.0015	-.0001	.1376	.0011	19.99
#4	-.0010	.0009	.1340	.0013	19.74
#5	-.0022	.0000	.1400	.0001	20.43

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.989
Stddev	.337
%RSD	2.1100

#1	15.683
#2	15.575
#3	16.147
#4	16.334
#5	16.205

Method: ISOU

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 02/25/99 14:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0013	-.0008	-.0007	-.0013	-.0005
Stddev	.0012	.0005	.0007	.0004	.0009
%RSD	98.89	60.73	99.60	34.43	175.5

#1	-.0018	-.0005	-.0003	-.0005	-.0006
#2	-.0026	-.0014	-.0010	-.0013	-.0015
#3	-.0016	-.0003	-.0009	-.0015	-.0004
#4	.0007	-.0006	-.0016	-.0016	-.0008
#5	-.0010	-.0013	.0003	-.0014	.0009

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.872
Stddev	.304
%RSD	1.9169

#1	15.756
#2	16.324
#3	15.820
#4	15.962
#5	15.496

HNF-1663 REV. 0

Method: ISOU

Sample Name: std-prep

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 14:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0018	.0026	.3434	.0043	47.00
Stddev	.0006	.0007	.0102	.0008	2.09
%RSD	33.18	25.65	2.974	18.45	4.440
#1	-.0014	.0036	.3275	.0042	45.11
#2	-.0023	.0018	.3390	.0041	49.11
#3	-.0022	.0026	.3488	.0032	44.47
#4	-.0019	.0027	.3499	.0054	48.55
#5	-.0009	.0022	.3520	.0044	47.76

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	16.827
Stddev	.338
%RSD	2.0116
#1	17.323
#2	16.593
#3	16.955
#4	16.451
#5	16.812

HNF-1663 REV. 0

Method: ISOU

Sample Name: blk-prep

Operator: kjt

Comment: 11 df

Run Time: 02/25/99 15:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0017	-.0018	-.0019	-.0027	.0021
Stddev	.0005	.0028	.0028	.0030	.0059
%RSD	32.07	153.6	142.4	113.5	287.2

#1	-.0018	.0002	-.0005	-.0003	.0057
#2	-.0011	-.0003	-.0006	-.0013	.0049
#3	-.0016	-.0014	-.0008	-.0020	.0037
#4	-.0013	-.0009	-.0009	-.0018	.0044
#5	-.0025	-.0067	-.0069	-.0079	-.0084

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.162
Stddev	1.840
%RSD	9.6047

#1	17.770
#2	18.611
#3	18.199
#4	18.858
#5	22.370

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002713

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0029	.0041	.6185	.0169	89.72
Stddev	.0010	.0015	.0147	.0007	2.14
%RSD	35.16	35.44	2.382	3.930	2.383
#1	.0040	.0059	.5975	.0160	88.19
#2	.0026	.0028	.6283	.0172	92.17
#3	.0025	.0030	.6360	.0169	88.60
#4	.0039	.0035	.6135	.0178	87.73
#5	.0016	.0056	.6173	.0166	91.89

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.736
Stddev	.501
%RSD	3.1854

#1	16.491
#2	15.935
#3	15.260
#4	15.662
#5	15.333

Method: ISOU   Sample Name: s98t002713_d   Operator: kjt
Comment: 201 df
Run Time: 02/25/99 15:08 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0017	.0074	.7876	.0215	112.9
Stddev	.0011	.0026	.0451	.0017	9.7
%RSD	68.32	35.58	5.731	7.896	8.581
#1	.0032	.0117	.8372	.0239	124.6
#2	.0015	.0052	.7891	.0213	116.5
#3	.0024	.0077	.8158	.0224	116.8
#4	.0003	.0054	.7184	.0204	99.78
#5	.0010	.0069	.7776	.0195	106.8

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.168
Stddev	1.513
%RSD	11.493

#1	10.816
#2	12.852
#3	13.438
#4	14.907
#5	13.825

Method: ISOU   Sample Name: s98t002846   Operator: kjt
Comment: 201 df
Run Time: 02/25/99 15:12 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	.0023	.3659	.0075	56.08
Stddev	.0010	.0008	.0183	.0016	3.35
%RSD	167.2	36.54	5.012	21.51	5.983
#1	.0003	.0016	.3604	.0093	55.38
#2	-.0005	.0032	.3400	.0059	50.53
#3	.0009	.0024	.3739	.0059	57.45
#4	.0021	.0013	.3652	.0089	58.34
#5	.0003	.0029	.3899	.0074	58.68

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.327
Stddev	.205
%RSD	1.5393

#1	13.343
#2	13.491
#3	13.336
#4	12.983
#5	13.479

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002846_d

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0016	.0031	.3803	.0090	57.57
Stddev	.0011	.0020	.0418	.0022	6.71
%RSD	69.41	65.22	10.98	24.30	11.65
#1	.0036	.0046	.4517	.0111	69.01
#2	.0009	.0009	.3558	.0100	52.51
#3	.0015	.0056	.3836	.0083	57.78
#4	.0013	.0014	.3571	.0055	55.14
#5	.0009	.0029	.3533	.0100	53.41

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.459
Stddev	1.516
%RSD	11.265

#1	10.822
#2	13.961
#3	13.618
#4	14.361
#5	14.532

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002846_s

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0015	.0060	.7111	.0149	101.1
Stddev	.0006	.0010	.0200	.0015	1.6
%RSD	36.91	16.51	2.816	9.743	1.620
#1	.0012	.0066	.7291	.0161	103.2
#2	.0012	.0045	.6804	.0137	99.16
#3	.0023	.0056	.7026	.0168	100.2
#4	.0010	.0065	.7262	.0140	102.4
#5	.0020	.0069	.7173	.0139	100.7

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.463
Stddev	.208
%RSD	1.3467

#1	15.260
#2	15.657
#3	15.319
#4	15.363
#5	15.716

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002846_a

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0038	.5048	.0110	73.43
Stddev	.0007	.0006	.0150	.0022	2.65
%RSD	78.51	14.91	2.977	19.65	3.604

#1	.0021	.0043	.5127	.0142	77.18
#2	.0005	.0029	.4969	.0118	74.49
#3	.0009	.0038	.5052	.0085	70.35
#4	.0007	.0041	.5244	.0102	71.59
#5	.0003	.0042	.4849	.0103	73.53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	16.701
Stddev	.614
%RSD	3.6789

#1	15.728
#2	16.513
#3	17.316
#4	16.972
#5	16.976

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002740

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0042	-.0010	.0324	-.0011	4.569
Stddev	.0013	.0009	.0010	.0004	.078
%RSD	31.64	95.29	3.114	36.52	1.702
#1	.0033	-.0016	.0313	-.0014	4.646
#2	.0029	-.0020	.0313	-.0014	4.557
#3	.0056	-.0003	.0332	-.0014	4.512
#4	.0036	-.0013	.0330	-.0009	4.652
#5	.0058	.0002	.0331	-.0005	4.480

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	18.760
Stddev	.276
%RSD	1.4734
#1	18.568
#2	18.467
#3	18.990
#4	18.667
#5	19.108

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t2740_d

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:32 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0060	-.0013	.0388	-.0007	5.581
Stddev	.0013	.0006	.0029	.0014	.452
%RSD	22.31	42.74	7.483	203.0	8.091
#1	.0069	-.0007	.0395	-.0021	5.836
#2	.0062	-.0008	.0393	-.0015	6.247
#3	.0059	-.0020	.0418	.0015	5.326
#4	.0072	-.0017	.0394	-.0008	5.149
#5	.0038	-.0014	.0340	-.0005	5.345

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.612
Stddev	1.097
%RSD	7.5054

#1	13.522
#2	13.800
#3	14.551
#4	14.874
#5	16.312

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t2740_s

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0049	.0012	.3831	.0051	50.99
Stddev	.0023	.0013	.0468	.0024	8.71
%RSD	47.05	101.1	12.21	46.51	17.07
#1	.0067	.0028	.4351	.0080	52.60
#2	.0074	.0019	.3888	.0024	62.36
#3	.0026	.0000	.3967	.0031	48.58
#4	.0051	-.0001	.3069	.0067	53.11
#5	.0024	.0016	.3877	.0051	38.31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.373
Stddev	1.708
%RSD	11.112

#1	13.191
#2	14.412
#3	17.439
#4	15.156
#5	16.668

Method: ISOU

Sample Name: s98t2740_a

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 15:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0058	.0010	.1618	.0020	22.24
Stddev	.0013	.0007	.0062	.0005	.77
%RSD	23.31	70.95	3.829	24.98	3.478

#1	.0052	.0020	.1594	.0017	21.84
#2	.0051	.0015	.1681	.0021	23.56
#3	.0080	.0006	.1630	.0027	21.86
#4	.0059	.0001	.1660	.0017	22.30
#5	.0046	.0010	.1524	.0015	21.65

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.827
Stddev	.403
%RSD	2.0331

#1	19.981
#2	19.335
#3	19.462
#4	20.133
#5	20.223

Method: ISOU

Sample Name: ccv

Operator: kjt

Comment:

Run Time: 02/25/99 15:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0013	.0006	.1377	.0014	19.29
Stddev	.0004	.0008	.0043	.0002	.54
%RSD	34.27	140.2	3.092	13.68	2.783

#1	-.0019	.0009	.1447	.0016	19.26
#2	-.0012	-.0005	.1342	.0014	18.95
#3	-.0015	.0017	.1374	.0017	19.55
#4	-.0010	.0001	.1343	.0012	18.64
#5	-.0008	.0007	.1381	.0012	20.02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.588
Stddev	.331
%RSD	1.6902

#1	19.390
#2	19.768
#3	19.565
#4	20.036
#5	19.180

HNF-1663 REV. 0

Method: ISOU

Sample Name: ccb

Operator: kjt

Comment: 2% hno3

Run Time: 02/25/99 15:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0015	-.0016	-.0005	-.0019	-.0007
Stddev	.0007	.0006	.0008	.0003	.0009
%RSD	43.93	35.50	158.4	17.50	125.0

#1	-.0008	-.0012	-.0019	-.0017	-.0007
#2	-.0019	-.0015	.0002	-.0015	.0008
#3	-.0020	-.0017	-.0005	-.0022	-.0015
#4	-.0022	-.0010	.0001	-.0021	-.0011
#5	-.0008	-.0025	-.0004	-.0023	-.0012

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	20.146
Stddev	.292
%RSD	1.4505

#1	19.960
#2	19.823
#3	20.341
#4	20.543
#5	20.063

LABCORE Completed Worklist Report for Worklist# 27706

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-6

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMS-U-QC U235	QC	0.000142	1.49e-4	104.930	% Recovery
1	ICV	0	EMS-U-QC U238	QC	0.020	2.20e-02	110.000	% Recovery
2	ICB	0	EMS-U-QC U233	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMS-U-QC U234	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMS-U-QC U235	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMS-U-QC U236	QC	1	<1.60e-5		ug/mL
2	ICB	0	EMS-U-QC U238	QC	1	<1.20e-5		ug/mL
3	STD-PREP	0	EMS-U-A1 U238-A1	SOLID	10	1.08e+01	108.000	% Recovery
4	BLNK-PREP	0	EMS-U-A1 U233-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMS-U-A1 U234-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMS-U-A1 U235-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMS-U-A1 U236-A1	SOLID	1	<1.60e-4		ug/g
4	BLNK-PREP	0	EMS-U-A1 U238-A1	SOLID	1	<1.20e-4		ug/g
5	SAMPLE	S98T002928	0 A EMS-U-A1 U233-A1	SOLID	N/A	< 4.918e-01	0.492	ug/g
5	SAMPLE	S98T002928	0 A EMS-U-A1 U234-A1	SOLID	N/A	< 4.918e-01	0.492	ug/g
5	SAMPLE	S98T002928	0 A EMS-U-A1 U235-A1	SOLID	N/A	< 4.918e-01	0.492	ug/g
5	SAMPLE	S98T002928	0 A EMS-U-A1 U236-A1	SOLID	N/A	< 6.558e-01	0.656	ug/g
5	SAMPLE	S98T002928	0 A EMS-U-A1 U238-A1	SOLID	N/A	< 3.674e+01	0.492	ug/g
6	DUP	S98T002928	0 A EMS-U-A1 U233-A1	SOLID	<4.92e-1	<4.58e-1		RPD
6	DUP	S98T002928	0 A EMS-U-A1 U234-A1	SOLID	<4.92e-1	<4.58e-1		RPD
6	DUP	S98T002928	0 A EMS-U-A1 U235-A1	SOLID	<4.92e-1	<4.58e-1		RPD
6	DUP	S98T002928	0 A EMS-U-A1 U236-A1	SOLID	<6.56e-1	<6.11e-1		RPD
6	DUP	S98T002928	0 A EMS-U-A1 U238-A1	SOLID	3.67e+01	3.65e+01	0.546	RPD
7	SPK	S98T002928	0 A EMS-U-A1 U235-A1	SOLID	14.22	1.49e+01	104.782	% Recovery
7	SPK	S98T002928	0 A EMS-U-A1 U238-A1	SOLID	2000	2.11e+03	105.500	% Recovery
8	SPK-POST	S98T002928	0 A EMS-U-A1 U235-A1	SOLID	0.0001422	1.53e-04	107.595	% Recovery
8	SPK-POST	S98T002928	0 A EMS-U-A1 U238-A1	SOLID	0.020	2.13e-02	106.500	% Recovery
9	SAMPLE	S98T002929	0 A EMS-U-A1 U233-A1	SOLID	N/A	< 4.744e-01	0.474	ug/g
9	SAMPLE	S98T002929	0 A EMS-U-A1 U234-A1	SOLID	N/A	< 4.744e-01	0.474	ug/g
9	SAMPLE	S98T002929	0 A EMS-U-A1 U235-A1	SOLID	N/A	< 4.744e-01	0.474	ug/g
9	SAMPLE	S98T002929	0 A EMS-U-A1 U236-A1	SOLID	N/A	< 6.326e-01	0.632	ug/g
9	SAMPLE	S98T002929	0 A EMS-U-A1 U238-A1	SOLID	N/A	< 2.211e+01	0.474	ug/g
10	DUP	S98T002929	0 A EMS-U-A1 U233-A1	SOLID	<4.74e-1	<4.90e-1		RPD
10	DUP	S98T002929	0 A EMS-U-A1 U234-A1	SOLID	<4.74e-1	<4.90e-1		RPD
10	DUP	S98T002929	0 A EMS-U-A1 U235-A1	SOLID	<4.74e-1	<4.90e-1		RPD
10	DUP	S98T002929	0 A EMS-U-A1 U236-A1	SOLID	<6.33e-1	<6.54e-1		RPD
10	DUP	S98T002929	0 A EMS-U-A1 U238-A1	SOLID	2.21e+01	2.36e+01	6.565	RPD
11	SPK-POST	S98T002929	0 A EMS-U-A1 U235-A1	SOLID	0.0001422	1.56e-04	109.705	% Recovery
11	SPK-POST	S98T002929	0 A EMS-U-A1 U238-A1	SOLID	0.020	2.21e-02	110.500	% Recovery
12	CCV	0	EMS-U-QC U235	QC	0.000142	1.52e-04	107.042	% Recovery
12	CCV	0	EMS-U-QC U238	QC	0.020	2.24e-02	112.000	% Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

2683

LABCORE Completed Worklist Report for Worklist# 27706

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
13	CCB	0			QMSU-QC U233	QC	1	<1.20e-5		ug/mL
13	CCB	0			QMSU-QC U234	QC	1	<1.20e-5		ug/mL
13	CCB	0			QMSU-QC U235	QC	1	<1.20e-5		ug/mL
13	CCB	0			QMSU-QC U236	QC	1	<1.60e-5		ug/mL
13	CCB	0			QMSU-QC U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 27706

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

03/03/99
_____
Reviewer Signature Date

03/03/99 09:43

ws2

LABCORE Data Entry Template for Worklist# 27706

Analyst: BJG Instrument: ICPMS1 _____ Book# 990226-6Method: LA-506-101 Rev/Mod A-Ø

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1 ICV				@MSU-QC	QC		
2 ICB				@MSU-QC	QC		
3 STD-PREP				@MSU-A1	SOLID		
4 BLNK-PREP				@MSU-A1	SOLID		
5 SAMPLE	S98T002928	0	A	@MSU-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
6 DUP	S98T002928	0	A	@MSU-A1	SOLID		
7 SPK	S98T002928	0	A	@MSU-A1	SOLID		
8 SPK-POST	S98T002928	0	A	@MSU-A1	SOLID		
9 SAMPLE	S98T002929	0	A	@MSU-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
10 DUP	S98T002929	0	A	@MSU-A1	SOLID		
11 SPK-POST	S98T002929	0	A	@MSU-A1	SOLID		
12 CCV				@MSU-QC	QC		
13 CCB				@MSU-QC	QC		

Final page for worklist # 27706

See Attached

Signature

Date

Signature

Date

Validated by
HAnast 3-3-99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2685

03/01/99 10:26
ws2

Page: 1

LABCORE Data Entry Template for Worklist# 27706

Analyst: B J G Instrument: ICPMS1 _____ Book# 890226-6Method: LA-506-101 Rev/Mod A-O

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 STD-PREP			@MSU-A1	SOLID		
4 BLNK-PREP			@MSU-A1	SOLID		
5 SAMPLE	S98T002928 0 A		@MSU-A1	SOLID	98000490 BY-105 (2)	
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1						
6 DUP	S98T002928 0 A		@MSU-A1	SOLID		
7 SPK	S98T002928 0 A		@MSU-A1	SOLID		
8 SPK-POST	S98T002928 0 A		@MSU-A1	SOLID		
9 SAMPLE	S98T002929 0 A		@MSU-A1	SOLID	98000490 BY-105 (2)	
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1						
10 DUP	S98T002929 0 A		@MSU-A1	SOLID		
11 SPK	S98T002929 0 A		@MSU-A1	SOLID		
~~12 SPK-POST~~	~~S98T002929 0 A~~		~~@MSU-A1~~	~~SOLID~~		
13 CCV			@MSU-QC	QC		
14 CCB			@MSU-QC	QC		
~~15 SAMPLE~~	~~S98T002930 0 A~~		~~@MSU-A1~~	~~SOLID~~	~~98000490 BY-105 (2)~~	
~~Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1~~						
~~16 DUP~~	~~S98T002930 0 A~~		~~@MSU-A1~~	~~SOLID~~		
~~17 SPK~~	~~S98T002930 0 A~~		~~@MSU-A1~~	~~SOLID~~		

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27706

S Type	Sample#	R A	Test	Matrix	Group#	Project
18 ~~SPK POST~~	~~S98T002930~~	~~0-A~~	@MSU-A1	SOLID		
19 CCV			@MSU-QC	QC		
20 CCB			@MSU-QC	QC		

Final page for worklist # 27706

[Signature] 3-1-99
Signature Date

Signature Date

PREP STD	.05-10-1	DF
PREP BLK	1-1-9-1	201
S98T00 2928	.05-10-1	10
2928-D		201
2928-A		
2928-S		
S98T00 2929		
2929-D		
2929-A		
S98T00 2930	.05-10-1-1-1-9-1	2010
2930-D		
2930-A		
13-1-99 2930-S		

POST SPIKE 2930-A = .02-1 of 10-11-95 (30 ppb)

SCANS = 982928d
982929d
982930d 3-1-99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2687

Analysis Report

03/01/99 14:21:23

page 1

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/01/99

14:20 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	221000.
Stddev	1499.
%RSD	.6782

#1	219900.
#2	219900.
#3	222300.
#4	219500.
#5	220900.
#6	218700.
#7	221900.
#8	220800.
#9	222900.
#10	223000.

3-1-99

B4-105 AND DIGEST

598700 2228-30

FILE # 00301A.TXT

W.L.# 22706

SCANS = 282928d

282929d

~~282930d~~ 3.1.58

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2688 TO 2707.

Method: ISOU Sample Name: icv
Comment: iso-u lcs
Run Time: 03/01/99 14:37 Type: Unk

Operator: bjg

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0015	.1491	.0023	21.95
Stddev	.0004	.0009	.0065	.0004	1.05
%RSD	4892.	58.72	4.361	18.48	4.785
#1	-.0004	.0004	.1381	.0019	20.12
#2	.0003	.0026	.1537	.0024	22.41
#3	.0005	.0022	.1542	.0026	22.46
#4	.0000	.0010	.1505	.0018	22.72
#5	-.0004	.0013	.1488	.0028	22.02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.008
Stddev	.629
%RSD	4.8337

#1	14.124
#2	12.678
#3	12.659
#4	12.716
#5	12.862

Method: ISOU

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 14:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0000	.0000	.0001	.0002
Stddev	.0005	.0003	.000	.0003	.0005
%RSD	411.7	697.6	1199.	221.5	204.3

#1	.0003	-.0002	.0003	-.0001	-.0002
#2	-.0002	.0001	-.0002	.0005	.0008
#3	.0005	.0005	.0001	-.0002	-.0003
#4	-.0009	-.0003	-.0003	.0004	.0004
#5	-.0003	.0001	.0000	.0001	.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.889
Stddev	.137
%RSD	1.0658

#1	13.009
#2	12.804
#3	13.003
#4	12.939
#5	12.693

Method: ISOU

Sample Name: prepsd

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 14:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	.0040	.3829	.0060	53.77
Stddev	.0003	.0005	.0039	.0006	.46
%RSD	106.5	13.65	1.015	10.52	.8462

#1	-.0005	.0038	.3834	.0060	54.23
#2	-.0003	.0035	.3784	.0053	53.86
#3	-.0005	.0036	.3888	.0056	53.01
#4	.0002	.0041	.3829	.0062	53.95
#5	-.0001	.0049	.3808	.0069	53.81

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.155
Stddev	.128
%RSD	.97218

#1	12.952
#2	13.158
#3	13.175
#4	13.184
#5	13.306

Method: ISOU

Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/01/99 15:00 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0001	.0000	.0001	.0065
Stddev	.001	.0003	.000	.0003	.0005
%RSD	9967.	366.1	1453.	301.3	8.312
#1	.0004	.0006	.0006	.0003	.0073
#2	.0006	-.0001	.0001	.0000	.0065
#3	-.0003	-.0001	-.0001	-.0003	.0059
#4	-.0004	.0001	-.0004	.0004	.0068
#5	-.0003	-.0001	-.0003	.0000	.0062

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.969
Stddev	.362
%RSD	2.7887

#1	12.329
#2	13.055
#3	13.129
#4	13.136
#5	13.198

Method: ISOU

Sample Name: s98t002928

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	.0003	.0057	.0004	.8965
Stddev	.0003	.0003	.0007	.0006	.0561
%RSD	62.12	99.86	12.63	173.3	6.255
#1	.0006	.0000	.0044	-.0002	.7991
#2	.0005	.0005	.0061	.0003	.9095
#3	.0011	.0006	.0060	-.0003	.9091
#4	.0005	.0002	.0061	.0008	.9233
#5	.0001	.0001	.0060	.0012	.9417

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.723
Stddev	.946
%RSD	6.8970

#1	15.391
#2	13.473
#3	13.439
#4	13.243
#5	13.069

Method: ISOU   Sample Name: s98t002928_d   Operator: bjg
Comment: 201 df
Run Time: 03/01/99 15:04 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0010	.0002	.0067	.0002	.9547
Stddev	.0005	.0003	.0010	.0004	.0322
%RSD	52.16	153.1	14.67	147.3	3.375

#1	.0004	-.0002	.0063	-.0003	.9007
#2	.0012	.0006	.0079	.0005	.9776
#3	.0016	.0002	.0057	.0005	.9802
#4	.0012	.0004	.0075	.0000	.9625
#5	.0005	.0001	.0058	.0005	.9526

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.987
Stddev	.326
%RSD	2.5098

#1	13.462
#2	12.995
#3	12.619
#4	12.760
#5	13.100

Method: ISOU

Sample Name: s98t002928_a

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0010	.0011	.1530	.0026	22.18
Stddev	.0002	.0002	.0018	.0005	.31
%RSD	20.35	14.91	1.187	20.24	1.415
#1	.0009	.0011	.1512	.0021	22.38
#2	.0008	.0008	.1515	.0033	22.25
#3	.0010	.0012	.1557	.0029	22.53
#4	.0008	.0013	.1529	.0022	21.74
#5	.0013	.0012	.1537	.0023	22.00

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.091
Stddev	.158
%RSD	1.2095

#1	12.930
#2	13.154
#3	12.913
#4	13.239
#5	13.220

Method: ISOU

Sample Name: s98t002928_s

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:10 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0033	.3842	.0056	55.34
Stddev	.0001	.0007	.0054	.0005	.71
%RSD	21.53	21.01	1.411	9.838	1.283
#1	.0009	.0026	.3882	.0061	55.63
#2	.0006	.0032	.3768	.0048	54.45
#3	.0006	.0032	.3905	.0060	55.98
#4	.0008	.0032	.3816	.0052	54.71
#5	.0005	.0045	.3836	.0057	55.92

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.138
Stddev	.136
%RSD	1.0351

#1	13.125
#2	13.277
#3	12.947
#4	13.260
#5	13.081

Method: ISOU

Sample Name: s98t002929

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	-.0001	.0034	.0000	.5592
Stddev	.0005	.0002	.0007	.000	.0079
%RSD	65.95	124.8	20.82	2292.	1.421

#1	.0003	-.0002	.0037	.0001	.5639
#2	.0013	-.0003	.0033	-.0006	.5603
#3	.0012	.0001	.0040	.0000	.5524
#4	.0004	-.0001	.0036	.0001	.5501
#5	.0005	-.0003	.0022	.0004	.5692

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.988
Stddev	.052
%RSD	.40088

#1	13.037
#2	12.964
#3	13.046
#4	12.969
#5	12.924

Method: ISOU

Sample Name: s98t002929_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	-.0002	.0036	-.0001	.5766
Stddev	.0004	.0005	.0004	.0001	.0175
%RSD	50.97	289.8	12.20	216.2	3.036
#1	.0004	-.0004	.0035	-.0001	.5486
#2	.0004	.0006	.0042	-.0002	.5820
#3	.0007	-.0003	.0039	.0000	.5714
#4	.0012	-.0001	.0031	.0001	.5911
#5	.0010	-.0007	.0034	-.0001	.5898

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.521
Stddev	.237
%RSD	1.7539

#1	13.915
#2	13.334
#3	13.528
#4	13.489
#5	13.338

Method: ISOU

Sample Name: s98t002929_a

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 15:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0010	.0012	.1562	.0023	22.62
Stddev	.0006	.0005	.0031	.0002	.56
%RSD	57.59	42.02	1.959	7.702	2.461
#1	.0012	.0012	.1547	.0024	22.47
#2	.0006	.0010	.1556	.0024	22.75
#3	.0015	.0014	.1547	.0021	22.17
#4	.0002	.0006	.1544	.0023	22.18
#5	.0016	.0020	.1616	.0025	23.51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.386
Stddev	.189
%RSD	1.4102

#1	13.398
#2	13.370
#3	13.489
#4	13.589
#5	13.085

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/01/99 15:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0015	.1518	.0020	22.42
Stddev	.000	.0004	.0030	.0004	.32
%RSD	1038.	30.25	1.948	19.92	1.449
#1	.0003	.0009	.1473	.0018	21.90
#2	.0000	.0020	.1512	.0022	22.60
#3	-.0002	.0013	.1540	.0015	22.60
#4	.0002	.0018	.1549	.0023	22.69
#5	-.0005	.0012	.1516	.0025	22.29

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.463
Stddev	.246
%RSD	1.8277

#1	13.889
#2	13.290
#3	13.301
#4	13.418
#5	13.414

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hmc3

Run Time: 03/01/99 15:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	-.0002	.0001	-.0001	.0003
Stddev	.0004	.0006	.0003	.0003	.0003
%RSD	470.1	227.3	280.8	284.4	131.5

#1	-.0004	.0003	-.0001	.0002	.0006
#2	-.0003	-.0004	.0002	.0000	.0001
#3	.0002	.0002	.0005	-.0006	-.0002
#4	.0003	-.0011	-.0003	.0000	.0006
#5	.0007	-.0002	.0003	-.0001	.0002

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.115
Stddev	.167
%RSD	1.2721

#1	12.976
#2	13.048
#3	13.261
#4	13.324
#5	12.965

2701

HNF-1663 REV. 0.

Method: ISOU

Sample Name: s98t002930

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 15:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0031	.3078	.0082	48.04
Stddev	.0004	.0006	.0069	.0005	1.23
%RSD	249.1	18.25	2.228	6.056	2.549
#1	.0003	.0029	.3081	.0077	47.14
#2	.0006	.0033	.2978	.0087	46.89
#3	.0001	.0027	.3171	.0076	49.07
#4	.0000	.0039	.3091	.0086	49.62
#5	-.0003	.0026	.3071	.0084	47.49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.364
Stddev	.247
%RSD	1.8466

#1	13.517
#2	13.582
#3	13.184
#4	13.021
#5	13.516

2702

Method: ISOU

Sample Name: s98t002930_d

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 15:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0017	.2068	.0053	31.90
Stddev	.0003	.0005	.0046	.0008	1.28
%RSD	641.2	30.44	2.205	15.90	4.000
#1	-.0001	.0012	.2042	.0046	31.12
#2	.0001	.0016	.2026	.0043	31.09
#3	.0002	.0026	.2088	.0052	32.97
#4	-.0003	.0015	.2045	.0059	30.75
#5	.0004	.0017	.2138	.0063	33.57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.944
Stddev	.386
%RSD	2.9782

#1	13.340
#2	13.252
#3	12.506
#4	13.054
#5	12.571

Method: ISOU

Sample Name: s98t002930_a

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 15:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0044	.4724	.0103	70.14
Stddev	.0003	.0002	.0114	.0003	1.53
%RSD	690.5	4.754	2.421	3.301	2.179
#1	.0002	.0042	.4920	.0102	72.70
#2	-.0003	.0046	.4638	.0106	69.17
#3	.0001	.0043	.4654	.0106	68.94
#4	.0005	.0047	.4684	.0104	70.36
#5	-.0002	.0044	.4724	.0098	69.55

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.221
Stddev	.228
%RSD	1.7247

#1	12.855
#2	13.479
#3	13.251
#4	13.216
#5	13.304

Method: ISOU

Sample Name: s98t002930_s

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 15:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0023	.2322	.0063	35.35
Stddev	.0006	.0003	.0064	.0005	1.05
%RSD	409.0	11.31	2.751	7.403	2.981

#1	.0002	.0027	.2248	.0061	33.49
#2	.0008	.0023	.2344	.0068	35.74
#3	-.0007	.0021	.2400	.0067	35.65
#4	.0004	.0022	.2265	.0062	35.70
#5	-.0001	.0022	.2353	.0057	36.14

Int. Std.	Ir [193]
Line	193/pulse
Units	Cts/S
Avg	13.251
Stddev	.338
%RSD	2.5474

#1	13.850
#2	13.055
#3	13.172
#4	13.089
#5	13.089

HNF-1663 REV. 0

Method: ISOU

Sample Name: ccv

Operator: bjg

Comment: iso-u los

Run Time: 03/01/99 15:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0013	.1557	.0020	22.41
Stddev	.0004	.0004	.0054	.0004	.60
%RSD	587.1	28.64	3.468	22.27	2.688
#1	.0001	.0011	.1472	.0015	21.56
#2	.0007	.0012	.1590	.0019	22.96
#3	.0002	.0008	.1570	.0023	22.35
#4	.0002	.0016	.1612	.0017	23.02
#5	.0005	.0016	.1541	.0026	22.18

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.231
Stddev	.496
%RSD	3.7460

#1	14.036
#2	12.957
#3	13.273
#4	12.728
#5	13.159

2706

Method: ISOU

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 15:51 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	.0001	-.0004	.0000	.0003
Stddev	.0003	.0002	.0002	.0003	.0002
%RSD	194.6	207.2	56.80	1433.	52.60

#1	.0000	.0001	-.0005	.0003	.0005
#2	-.0002	.0001	-.0003	.0004	.0003
#3	.0001	.0003	-.0007	.0001	.0005
#4	-.0006	-.0001	-.0001	-.0002	.0001
#5	.0000	-.0001	-.0004	-.0004	.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.017
Stddev	.178
%RSD	1.3659

#1	13.123
#2	12.723
#3	12.971
#4	13.121
#5	13.144

2707

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 9 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 2708 THRU 3069)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INDUCTIVELY COUPLED PLASMA/MASS SPECTROMETER ANALYSIS, ICP/MS WORKLIST 27707 THRU 27946		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17Sec 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Sec 6
of 17

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 7
of 17

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 27707

8/03/04/99

Analyst: bjg

Instrument: ICPMS1

Book#: B559 8.B59 spike

Method: LA-506-101 Rev/Mod A-D

990303-3 III, CCO

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	MSACT TH232	QC	0.50e-2	5.18e-03	103.600 % Recovery	
1 ICV	0	MSACT NP237	QC	4.94e-3	5.90e-03	119.433 % Recovery	
1 ICV	0	MSACT PU239	QC	0.502e-2	5.14e-03	102.390 % Recovery	
1 ICV	0	MSACT 241	QC	0.513e-2	5.53e-03	107.797 % Recovery	
2 ICB	0	MSAC-QC PA231	QC	1	<5.57e-6	ug/mL	
2 ICB	0	MSAC-QC TH229	QC	1	<4.46e-6	ug/mL	
2 ICB	0	MSAC-QC TH230	QC	1	<3.55e-6	ug/mL	
2 ICB	0	MSAC-QC TH232	QC	1	<1.93e-6	ug/mL	
2 ICB	0	MSAC-QC NP237	QC	1	<4.67e-6	ug/mL	
2 ICB	0	MSAC-QC PU239	QC	1	<3.47e-6	ug/mL	
2 ICB	0	MSAC-QC PU240	QC	1	<3.55e-6	ug/mL	
2 ICB	0	MSAC-QC PU242	QC	1	<2.78e-6	ug/mL	
2 ICB	0	MSAC-QC 241	QC	1	<3.71e-6	ug/mL	
2 ICB	0	MSAC-QC 243	QC	1	<3.83e-6	ug/mL	
2 ICB	0	MSAC-QC 244	QC	1	<3.41e-6	ug/mL	
3 BLNK-PREP	0	MSAC-A1 PA231-A1 SOLID		1	<5.57e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 TH229-A1 SOLID		1	<4.46e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 TH230-A1 SOLID		1	<3.55e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 TH232-A1 SOLID		1	2.12e-05	2.120e-005 ug/g	
3 BLNK-PREP	0	MSAC-A1 NP237-A1 SOLID		1	<4.67e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 PU239-A1 SOLID		1	<3.47e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 PU240-A1 SOLID		1	<3.55e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 PU242-A1 SOLID		1	<2.78e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 241-A1 SOLID		1	<3.71e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 243-A1 SOLID		1	<3.83e-5	ug/g	
3 BLNK-PREP	0	MSAC-A1 244-A1 SOLID		1	<3.41e-5	ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 PA231-A1 SOLID		N/A <	2.200e-01	0.220 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 TH229-A1 SOLID		N/A <	1.763e-01	0.176 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 TH230-A1 SOLID		N/A <	1.403e-01	0.140 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 TH232-A1 SOLID		N/A	1.093e+00	7.63e-002 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 NP237-A1 SOLID		N/A <	1.845e-01	0.184 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 PU239-A1 SOLID		N/A <	1.371e-01	0.137 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 PU240-A1 SOLID		N/A <	1.404e-01	0.140 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 PU242-A1 SOLID		N/A <	1.098e-01	0.110 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 241-A1 SOLID		N/A <	1.464e-01	0.146 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 243-A1 SOLID		N/A <	1.515e-01	0.151 ug/g	
4 SAMPLE	S98T003394 0 A	MSAC-A1 244-A1 SOLID		N/A <	1.349e-01	0.135 ug/g	
5 DUP	S98T003394 0 A	MSAC-A1 PA231-A1 SOLID		<2.20e-1	<2.30e-1	RFD	
5 DUP	S98T003394 0 A	MSAC-A1 TH229-A1 SOLID		<1.76e-1	<1.84e-1	RFD	
5 DUP	S98T003394 0 A	MSAC-A1 TH230-A1 SOLID		<1.40e-1	<1.47e-1	RFD	
5 DUP	S98T003394 0 A	MSAC-A1 TH232-A1 SOLID		1.09e+00	1.22e+00	11.255 RFD	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27707

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 DUP	S98T003394	0 A	OMSAC-A1 NP237-A1	SOLID	<1.84e-1	<1.93e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 PU239-A1	SOLID	<1.37e-1	<1.43e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 PU240-A1	SOLID	<1.40e-1	<1.47e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 PU242-A1	SOLID	<1.10e-1	<1.15e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 241-A1	SOLID	<1.46e-1	<1.53e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 243-A1	SOLID	<1.51e-1	<1.58e-1		RPD
5 DUP	S98T003394	0 A	OMSAC-A1 244-A1	SOLID	<1.35e-1	<1.41e-1		RPD
6 SPK-POST	S98T003394	0 A	OMSAC-A1 TH232-A1	SOLID	0.020	1.99e-02	99.500 % Recovery	
6 SPK-POST	S98T003394	0 A	OMSAC-A1 NP237-A1	SOLID	0.01976	2.32e-02	117.409 % Recovery	
6 SPK-POST	S98T003394	0 A	OMSAC-A1 PU239-A1	SOLID	0.02008	2.03e-02	102.092 % Recovery	
6 SPK-POST	S98T003394	0 A	OMSAC-A1 241-A1	SOLID	0.02052	2.10e-02	106.238 % Recovery	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 PA231-A1	SOLID	N/A	< 2.268e+01	22.680 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 TH229-A1	SOLID	N/A	< 1.817e+01	18.170 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 TH230-A1	SOLID	N/A	< 1.446e+01	14.460 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 TH232-A1	SOLID	N/A	< 7.859e+00	7.859 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 NP237-A1	SOLID	N/A	< 1.901e+01	19.010 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 PU239-A1	SOLID	N/A	< 1.413e+01	14.130 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 PU240-A1	SOLID	N/A	< 1.447e+01	14.470 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 PU242-A1	SOLID	N/A	< 1.132e+01	11.320 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 241-A1	SOLID	N/A	< 1.509e+01	15.090 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 243-A1	SOLID	N/A	< 1.561e+01	15.610 ug/g	
7 SAMPLE	S98T003399	0 A	OMSAC-A1 244-A1	SOLID	N/A	< 1.390e+01	13.900 ug/g	
8 DUP	S98T003399	0 A	OMSAC-A1 PA231-A1	SOLID	<2.27e1	<2.17e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 TH229-A1	SOLID	<1.82e1	<1.74e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 TH230-A1	SOLID	<1.45e1	<1.38e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 TH232-A1	SOLID	<7.86e0	<7.52e0		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 NP237-A1	SOLID	<1.90e1	<1.82e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 PU239-A1	SOLID	<1.41e1	<1.35e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 PU240-A1	SOLID	<1.45e1	<1.38e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 PU242-A1	SOLID	<1.13e1	<1.08e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 241-A1	SOLID	<1.51e1	<1.44e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 243-A1	SOLID	<1.56e1	<1.49e1		RPD
8 DUP	S98T003399	0 A	OMSAC-A1 244-A1	SOLID	<1.39e1	<1.33e1		RPD
9 SPK-POST	S98T003399	0 A	OMSAC-A1 TH232-A1	SOLID	0.020	1.99e-02	99.500 % Recovery	
9 SPK-POST	S98T003399	0 A	OMSAC-A1 NP237-A1	SOLID	0.01976	2.31e-02	116.903 % Recovery	
9 SPK-POST	S98T003399	0 A	OMSAC-A1 PU239-A1	SOLID	0.02008	2.03e-02	101.096 % Recovery	
9 SPK-POST	S98T003399	0 A	OMSAC-A1 241-A1	SOLID	0.02052	2.19e-02	106.725 % Recovery	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 PA231-A1	SOLID	N/A	< 2.295e-01	0.230 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 TH229-A1	SOLID	N/A	< 1.839e-01	0.184 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 TH230-A1	SOLID	N/A	< 1.463e-01	0.146 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 TH232-A1	SOLID	N/A	1.954e+00	7.95e-002 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 NP237-A1	SOLID	N/A	< 1.924e-01	0.192 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 PU239-A1	SOLID	N/A	< 1.430e-01	0.143 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 PU240-A1	SOLID	N/A	< 1.464e-01	0.146 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 PU242-A1	SOLID	N/A	< 1.145e-01	0.115 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 241-A1	SOLID	N/A	< 1.527e-01	0.153 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 243-A1	SOLID	N/A	< 1.580e-01	0.158 ug/g	
10 SAMPLE	S98T003389	0 A	OMSAC-A1 244-A1	SOLID	N/A	< 1.407e-01	0.141 ug/g	
11 DUP	S98T003389	0 A	OMSAC-A1 PA231-A1	SOLID	<2.30e-1	<2.12e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27707

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11 DUP	S98T003389	0 A	QMSAC-A1 TH229-A1	SOLID	<1.84e-1	<1.70e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 TH230-A1	SOLID	<1.46e-1	<1.35e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 TH232-A1	SOLID	1.95e+00	2.10e+00	7.407	RPD
11 DUP	S98T003389	0 A	QMSAC-A1 NP237-A1	SOLID	<1.92e-1	<1.78e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 PU239-A1	SOLID	<1.43e-1	<1.32e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 PU240-A1	SOLID	<1.46e-1	<1.35e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 PU242-A1	SOLID	<1.15e-1	<1.06e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 241-A1	SOLID	<1.53e-1	<1.41e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 243-A1	SOLID	<1.58e-1	<1.46e-1		RPD
11 DUP	S98T003389	0 A	QMSAC-A1 244-A1	SOLID	<1.41e-1	<1.30e-1		RPD
12 SPK-POST	S98T003389	0 A	QMSAC-A1 TH232-A1	SOLID	0.020	1.95e-02	97.500 %	Recovery
12 SPK-POST	S98T003389	0 A	QMSAC-A1 NP237-A1	SOLID	0.01976	2.28e-02	115.385 %	Recovery
12 SPK-POST	S98T003389	0 A	QMSAC-A1 PU239-A1	SOLID	0.02008	2.00e-02	99.602 %	Recovery
12 SPK-POST	S98T003389	0 A	QMSAC-A1 241-A1	SOLID	0.02052	2.14e-02	104.288 %	Recovery
13 CCV	0		QMSACT TH232	QC	0.50e-2	5.15e-03	103.000 %	Recovery
13 CCV	0		QMSACT NP237	QC	4.94e-3	5.98e-03	121.053 %	Recovery
13 CCV	0		QMSACT PU239	QC	0.502e-2	5.19e-03	103.386 %	Recovery
13 CCV	0		QMSACT 241	QC	0.513e-2	5.67e-03	110.526 %	Recovery
14 CCB	0		QMSAC-QC PA231	QC	1	<5.57e-6		ug/mL
14 CCB	0		QMSAC-QC TH229	QC	1	<4.46e-6		ug/mL
14 CCB	0		QMSAC-QC TH230	QC	1	<3.55e-6		ug/mL
14 CCB	0		QMSAC-QC TH232	QC	1	<1.93e-6		ug/mL
14 CCB	0		QMSAC-QC NP237	QC	1	<4.67e-6		ug/mL
14 CCB	0		QMSAC-QC PU239	QC	1	<3.47e-6		ug/mL
14 CCB	0		QMSAC-QC PU240	QC	1	<3.55e-6		ug/mL
14 CCB	0		QMSAC-QC PU242	QC	1	<2.78e-6		ug/mL
14 CCB	0		QMSAC-QC 241	QC	1	<3.71e-6		ug/mL
14 CCB	0		QMSAC-QC 243	QC	1	<3.83e-6		ug/mL
14 CCB	0		QMSAC-QC 244	QC	1	<3.41e-6		ug/mL

Final page for worklist# 27707

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sand M. Pang 03/09/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27707

Analyst: BTG Instrument: ICPMS1 _____ Book# 8857 SPILL, 990803-8 ICP/CCV

Method: LA-506-101 Rev/Mod A0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC		
2 ICB			@MSACTA1	SOLID		
3 BLINK-PREP			@MSAC-A1	SOLID		
4 SAMPLE	S98T003394	0 A	@MSAC-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
5 DUP	S98T003394	0 A	@MSAC-A1	SOLID		
6 SPK-POST	S98T003394	0 A	@MSAC-A1	SOLID		
7 SAMPLE	S98T003399	0 A	@MSAC-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
8 DUP	S98T003399	0 A	@MSAC-A1	SOLID		
9 SPK-POST	S98T003399	0 A	@MSAC-A1	SOLID		
10 SAMPLE	S98T003389	0 A	@MSAC-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
11 DUP	S98T003389	0 A	@MSAC-A1	SOLID		
12 SPK-POST	S98T003389	0 A	@MSAC-A1	SOLID		
13 CCV			@MSACT	QC		
14 CCB			@MSACTA1	SOLID		

Data Entry Comments:

Uploaded + Validated by:

Sal M. Perry

03/04/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2711

12/29/98 09:18

A-0004-1

Page: 2

LABCORE Data Entry Template for Worklist# 27707

S Type	Sample#	R A Test	Matrix	Group#	Project

Final page for worklist # 27707

Analyst Signature *3-3-99*

Analyst Signature *Date*

PREP BLK 1-1-9-1
598700 3384 .05-1-10-1
3384-D
3384-A
↓
598700 3389 .05-1-10-1
3389-D
3389-A
↓
598700 3387 .05-1-10-1
3387-D
3387-A
↓

D.F.

10

201

↓

20, 301

↓

201

↓

POST SPICES = .04 ml of ISART SPICE 81859 in 20-1
ICU/CCU = 11859 DILUTED 1:1 (990803-3)
PREVIOUSLY SAMPLED

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2712

Analysis Report

03/03/99 13:54:20

page 1

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/03/99 13:53 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	218200.
Stddev	2140.
%RSD	.9807

#1	220700.
#2	220400.
#3	219700.
#4	215000.
#5	218900.
#6	215800.
#7	215500.
#8	219800.
#9	217200.
#10	219000.

ACTINIDE
BY-105 558700 3384 28, 28

ACID DIGEST

FILE # AC0303A.TPT

WL # 22207

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2713 TO 2747

Method: ACTINIDE Sample Name: cal std
 Comment: 0.010 ppm cal
 Run Time: 03/03/99 14:16 Type: Unk

Operator: bjg

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4097	.5040	794.7	877.3	1.555
Stddev	.0505	.0462	18.5	9.6	.110
%RSD	12.31	9.160	2.327	1.095	7.080

#1	.4394	.4939	766.3	861.8	1.379
#2	.3760	.5607	787.7	878.5	1.628
#3	.3380	.4940	803.7	880.9	1.518
#4	.4564	.5330	814.4	888.0	1.650
#5	.4386	.4386	801.2	877.2	1.598

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.888	.6642	.3985	.3910	.4329
Stddev	.088	.0724	.0535	.0200	.0533
%RSD	4.668	10.90	13.44	5.103	12.31

#1	1.761	.6948	.3747	.3747	.3713
#2	1.937	.7520	.4734	.4129	.5137
#3	1.914	.5817	.3802	.3932	.4420
#4	1.842	.5969	.3351	.3670	.4022
#5	1.986	.6956	.4292	.4073	.4355

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4262	.3980
Stddev	.0190	.0528
%RSD	4.452	13.28

#1	.4326	.4292
#2	.4197	.3659
#3	.3965	.4745
#4	.4373	.3447
#5	.4449	.3760

2714

Sample Name: cal std Run Time: 03/03/99 14:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.11
Stddev	3.53
%RSD	3.4599
#1	97.864
#2	99.289
#3	102.57
#4	104.44
#5	106.39

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 03/03/99 14:21 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4752	1.267	7976.	8449.	12.47
Stddev	.0374	.060	104.	154.	.35
%RSD	7.878	4.750	1.308	1.824	2.830
#1	.4941	1.169	7938.	8238.	12.12
#2	.4243	1.292	8088.	8417.	13.00
#3	.5184	1.331	7815.	8390.	12.63
#4	.4886	1.270	8028.	8604.	12.37
#5	.4505	1.275	8009.	8596.	12.24

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	15.03	2.899	.4208	.4205	.4226
Stddev	.45	.105	.0242	.0355	.0317
%RSD	2.988	3.624	5.763	8.447	7.492
#1	14.69	2.746	.4339	.3769	.4751
#2	15.59	2.996	.3919	.4534	.4049
#3	14.84	2.998	.4480	.4288	.4096
#4	15.42	2.893	.3986	.4533	.3954
#5	14.59	2.861	.4314	.3902	.4283

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3922	.3839
Stddev	.0233	.0302
%RSD	5.949	7.861
#1	.3927	.3611
#2	.4081	.4016
#3	.4000	.4288
#4	.4082	.3632
#5	.3521	.3648

Sample Name: cal std Run Time: 03/03/99 14:21

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.22
Stddev	.97
%RSD	.93083

#1	105.24
#2	102.92
#3	104.17
#4	103.70
#5	105.08

Method: ACTINIDE Sample Name: cal blank

Operator: bjg

Comment: background

Run Time: 03/03/99 14:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3978	.3764	.6719	.4794	.3948
Stddev	.0203	.0527	.0490	.0164	.0301
%RSD	5.094	14.00	7.293	3.428	7.622

#1	.4066	.3818	.7170	.4811	.3694
#2	.3858	.3262	.6587	.4674	.4297
#3	.3696	.3289	.7204	.4730	.3853
#4	.4051	.4545	.6617	.5071	.4236
#5	.4218	.3907	.6016	.4683	.3659

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3831	.4097	.4209	.4183	.4065
Stddev	.0391	.0177	.0178	.0236	.0341
%RSD	10.22	4.316	4.238	5.641	8.384

#1	.3600	.4004	.4221	.4066	.3631
#2	.4360	.3952	.3952	.3890	.4203
#3	.4135	.4197	.4354	.4354	.3821
#4	.3587	.4360	.4391	.4484	.4174
#5	.3473	.3969	.4125	.4125	.4497

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3978	.3891
Stddev	.0378	.0421
%RSD	9.493	10.81

#1	.3725	.3942
#2	.4015	.3670
#3	.3571	.4041
#4	.4020	.3339
#5	.4559	.4466

2718

Sample Name: cal blank   Run Time: 03/03/99 14:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.07
Stddev	.69
%RSD	.64071

#1	107.39
#2	106.27
#3	106.42
#4	107.80
#5	107.49

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 14:41 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4307	.4204	412.2	7.519	.9788
Stddev	.0168	.0546	2.2	.092	.0237
%RSD	3.893	12.99	.5242	1.218	2.419

#1	.4264	.4393	413.3	7.535	1.013
#2	.4426	.3821	410.0	7.566	.9489
#3	.4470	.4856	410.5	7.577	.9873
#4	.4044	.4465	411.8	7.561	.9738
#5	.4332	.3485	415.3	7.358	.9706

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	510.0	458.2	12.75	.4669	.4237
Stddev	6.5	3.9	.28	.0274	.0271
%RSD	1.284	.8403	2.163	5.862	6.391

#1	507.8	453.1	12.39	.4489	.3944
#2	502.3	460.3	12.77	.4362	.4235
#3	509.7	455.0	13.09	.4888	.4599
#4	510.0	461.3	12.58	.5014	.4012
#5	520.4	461.2	12.92	.4592	.4397

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	508.2	.4295
Stddev	3.6	.0453
%RSD	.7176	10.54

#1	503.4	.4200
#2	508.6	.3853
#3	511.7	.5017
#4	505.7	.4400
#5	511.5	.4006

Sample Name: icv Run Time: 03/03/99 14:41

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	103.53
Stddev	.89
%RSD	.86132

#1	103.96
#2	104.68
#3	103.65
#4	103.03
#5	102.34

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 14:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3800	.3862	.5653	.4584	.4125
Stddev	.0182	.0517	.0452	.0385	.0558
%RSD	4.789	13.39	7.992	8.397	13.54

#1	.3867	.3867	.5556	.4062	.4809
#2	.4084	.3039	.4999	.4476	.3692
#3	.3719	.4136	.5547	.4810	.3431
#4	.3714	.3844	.6092	.5082	.4268
#5	.3617	.4425	.6072	.4489	.4425

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3896	.4453	.3503	.4143	.3956
Stddev	.0332	.0328	.0379	.0159	.0326
%RSD	8.513	7.371	10.81	3.827	8.245

#1	.3672	.4094	.3542	.4094	.3542
#2	.3921	.4215	.4019	.3953	.4051
#3	.4457	.4938	.2982	.4136	.3687
#4	.3779	.4463	.3356	.4137	.4268
#5	.3649	.4554	.3617	.4392	.4231

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4241	.4038
Stddev	.0273	.0423
%RSD	6.449	10.47

#1	.4159	.4387
#2	.4640	.3757
#3	.4265	.4585
#4	.3877	.3616
#5	.4263	.3843

2722

HNF-1663 REV. 0

Sample Name: icb Run Time: 03/03/99 14:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.82
Stddev	.77
%RSD	.75254

#1	102.58
#2	102.02
#3	103.96
#4	102.32
#5	103.21

2723

Method: ACTINIDE Sample Name: prepbkl

Operator: bjg

Comment: 10 df

Run Time: 03/03/99 14:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3592	.3662	.6412	.6015	.3715
Stddev	.0404	.0453	.0498	.0273	.0248
%RSD	11.24	12.36	7.771	4.535	6.677

#1	.3969	.4376	.7188	.6063	.3938
#2	.3830	.3498	.6575	.6002	.3951
#3	.3855	.3612	.6132	.5828	.3430
#4	.3146	.3691	.5899	.6443	.3479
#5	.3162	.3132	.6265	.5738	.3777

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3792	.5394	.3870	.3915	.3892
Stddev	.0331	.0453	.0448	.0388	.0376
%RSD	8.728	8.394	11.58	9.909	9.649

#1	.3532	.5594	.4094	.3625	.4094
#2	.3679	.5851	.3529	.3830	.3529
#3	.4371	.5707	.3734	.4310	.3642
#4	.3721	.4870	.4538	.4326	.4447
#5	.3659	.4948	.3455	.3484	.3747

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3826	.3998
Stddev	.0461	.0299
%RSD	12.04	7.485

#1	.4313	.4282
#2	.3589	.3830
#3	.4189	.3734
#4	.3176	.3781
#5	.3864	.4362

2724

Sample Name: prepblk Run Time: 03/03/99 14:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	110.21
Stddev	2.56
%RSD	2.3238

#1	106.65
#2	110.53
#3	109.82
#4	110.19
#5	113.86

2725

Method: ACTINIDE Sample Name: s98t003394

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U ₂ [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3747	.3668	2.670	44.72	.3556
Stddev	.0440	.0368	.068	.51	.0333
%RSD	11.75	10.03	2.534	1.141	9.370

#1	.3296	.3514	2.727	43.98	.3856
#2	.4273	.3635	2.605	45.03	.3571
#3	.3281	.4310	2.612	44.70	.3088
#4	.3972	.3418	2.754	44.56	.3386
#5	.3914	.3465	2.650	45.33	.3882

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4246	.4274	.3779	.4150	.3624
Stddev	.0344	.0432	.0316	.0325	.0331
%RSD	8.107	10.10	8.355	7.838	9.122

#1	.3887	.3700	.3482	.3700	.3576
#2	.4464	.3922	.3890	.3986	.3540
#3	.3988	.4535	.4181	.4535	.4085
#4	.4720	.4655	.3907	.4167	.3744
#5	.4170	.4555	.3432	.4363	.3176

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3761	.3575
Stddev	.0322	.0257
%RSD	8.570	7.183

#1	.3451	.3265
#2	.3699	.3603
#3	.3763	.3409
#4	.4297	.3939
#5	.3593	.3657

Sample Name: s98t003394 Run Time: 03/03/99 15:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.33
Stddev	1.78
%RSD	1.7089

#1	107.20
#2	104.53
#3	103.63
#4	102.39
#5	103.91

2727

Method: ACTINIDE

Sample Name: s98t003394_d

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:09 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3499	.3733	2.828	53.10	.3954
Stddev	.0541	.0455	.141	.94	.0297
%RSD	15.45	12.20	4.971	1.768	7.522

#1	.3204	.3302	2.652	54.34	.3728
#2	.2685	.3656	2.977	52.39	.3915
#3	.3893	.3827	2.800	52.29	.4449
#4	.3929	.3418	2.747	52.62	.3961
#5	.3782	.4461	2.964	53.86	.3718

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4692	.4837	.3732	.3765	.4213
Stddev	.0027	.0513	.0455	.0314	.0331
%RSD	.5805	10.60	12.19	8.339	7.850

#1	.4676	.4218	.3891	.3270	.4022
#2	.4659	.5144	.3850	.3753	.4400
#3	.4711	.4351	.3075	.4122	.4187
#4	.4727	.5334	.3545	.3769	.3801
#5	.4687	.5140	.4299	.3912	.4655

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3855	.3843
Stddev	.0503	.0276
%RSD	13.06	7.179

#1	.4153	.3989
#2	.3721	.4206
#3	.3860	.3468
#4	.4440	.3801
#5	.3103	.3750

Sample Name: s98t003394_d Run Time: 03/03/99 15:09

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.87
Stddev	1.01
%RSD	.98643

#1	101.94
#2	103.03
#3	101.89
#4	104.36
#5	103.11

Method: ACTINIDE Sample Name: s98t003394_a

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:15 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3691	.5330	1585.	48.98	2.613
Stddev	.0291	.0399	20.	.63	.082
%RSD	7.875	7.487	1.247	1.288	3.130

#1	.3311	.4740	1613.	49.62	2.623
#2	.3673	.5510	1562.	48.28	2.484
#3	.3961	.5156	1576.	48.40	2.654
#4	.3518	.5788	1595.	49.57	2.600
#5	.3991	.5457	1577.	49.03	2.703

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2000.	1822.	49.29	.6809	.4077
Stddev	27.	25.	.72	.0375	.0497
%RSD	1.326	1.349	1.466	5.503	12.19

#1	2034.	1843.	49.69	.6395	.3668
#2	1967.	1784.	48.11	.7377	.3860
#3	1984.	1829.	49.23	.6634	.4873
#4	2018.	1842.	50.01	.6683	.3741
#5	1996.	1812.	49.40	.6954	.4241

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2002.	.4366
Stddev	45.	.0243
%RSD	2.246	5.562

#1	2067.	.4350
#2	1957.	.4264
#3	1974.	.4590
#4	2028.	.4605
#5	1984.	.4022

Sample Name: s98t003394_a Run Time: 03/03/99 15:15

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	105.39
Stddev	1.89
%RSD	1.7889
#1	102.68
#2	107.09
#3	106.02
#4	104.24
#5	106.90

2731

Method: ACTINIDE Sample Name: s98t003399

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 15:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3903	.3597	.5623	851.3	.3853
Stddev	.0318	.0387	.0431	11.7	.0345
%RSD	8.148	10.75	7.672	1.374	8.946

#1	.4323	.3292	.5687	845.6	.3592
#2	.3482	.3551	.4882	865.7	.4029
#3	.3829	.3965	.5964	847.8	.4269
#4	.3793	.3161	.5889	836.9	.3959
#5	.4085	.4018	.5692	860.5	.3415

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.747	.9885	.3837	.3841	.3803
Stddev	.093	.0491	.0272	.0413	.0424
%RSD	5.302	4.963	7.078	10.76	11.14

#1	1.729	1.021	.3625	.3492	.3226
#2	1.762	.9286	.3517	.4199	.3551
#3	1.843	1.054	.4202	.4371	.3931
#4	1.600	.9714	.3926	.3626	.4325
#5	1.802	.9677	.3918	.3516	.3985

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3798	.3849
Stddev	.0403	.0219
%RSD	10.60	5.701

#1	.3858	.3991
#2	.3482	.3517
#3	.4473	.3795
#4	.3560	.4092
#5	.3616	.3851

2732

HNF-1663 REV. 0

Sample Name: s98t003399 Run Time: 03/03/99 15:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	99.196
Stddev	1.152
%RSD	1.1617

#1	100.23
#2	97.633
#3	98.372
#4	100.19
#5	99.547

2733

Method: ACTINIDE Sample Name: s98t003399_d

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 15:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3775	.3586	.5117	993.6	.3523
Stddev	.0375	.0289	.0298	15.1	.0491
%RSD	9.923	8.048	5.822	1.522	13.93

#1	.3575	.3976	.5179	996.8	.3709
#2	.4364	.3451	.5277	993.3	.2977
#3	.3634	.3501	.5167	970.5	.4267
#4	.3901	.3765	.5359	1013.	.3358
#5	.3402	.3235	.4603	994.7	.3302

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.026	1.033	.4082	.3733	.3919
Stddev	.034	.093	.0391	.0255	.0286
%RSD	1.699	9.051	9.566	6.828	7.286

#1	2.048	.9557	.3475	.3709	.4110
#2	2.074	.9878	.4059	.4026	.3484
#3	1.994	1.054	.4534	.3667	.4201
#4	1.998	1.187	.4274	.3358	.3799
#5	2.015	.9806	.4069	.3903	.4003

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3868	.4171
Stddev	.0430	.0310
%RSD	11.12	7.422

#1	.3976	.4010
#2	.3484	.4330
#3	.3367	.3701
#4	.4376	.4410
#5	.4136	.4403

2734

Sample Name: s98t003399_d Run Time: 03/03/99 15:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	99.296
Stddev	.825
%RSD	.83074

#1	99.753
#2	98.536
#3	99.984
#4	98.270
#5	99.935

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t003399_a

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 15:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4488	.5749	1582.	880.1	2.588
Stddev	.0409	.0333	32.	20.8	.155
%RSD	9.107	5.797	2.013	2.363	6.001
#1	.4917	.5948	1536.	859.3	2.389
#2	.4322	.5616	1600.	901.1	2.495
#3	.4506	.6220	1563.	871.6	2.572
#4	.4805	.5380	1600.	864.9	2.744
#5	.3892	.5579	1613.	903.6	2.741

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1993.	1812.	49.65	.6434	.3920
Stddev	38.	45.	1.16	.0352	.0361
%RSD	1.882	2.479	2.340	5.464	9.219
#1	1943.	1749.	48.51	.6114	.3721
#2	2031.	1844.	48.28	.6160	.3778
#3	1979.	1791.	50.20	.6792	.3900
#4	1983.	1812.	50.78	.6835	.3655
#5	2030.	1863.	50.45	.6268	.4546

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2010.	.4344
Stddev	34.	.0416
%RSD	1.691	9.572
#1	1965.	.3788
#2	2025.	.4595
#3	1999.	.4539
#4	2006.	.4771
#5	2057.	.4029

2736

Sample Name: s98t003399_a Run Time: 03/03/99 15:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	98.539
Stddev	1.319
%RSD	1.3385

#1	100.32
#2	97.940
#3	99.135
#4	98.508
#5	96.788

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t003389

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:47 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3513	.3383	4.239	183.4	.3681
Stddev	.0350	.0201	.118	5.0	.0168
%RSD	9.966	5.939	2.792	2.704	4.553
#1	.3654	.3363	4.194	176.5	.3622
#2	.3500	.3726	4.086	181.5	.3921
#3	.3993	.3214	4.204	185.5	.3636
#4	.3375	.3343	4.391	183.6	.3472
#5	.3043	.3269	4.318	189.9	.3755

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.6780	1.215	.3959	.3856	.4211
Stddev	.0650	.091	.0203	.0436	.0203
%RSD	9.590	7.485	5.116	11.30	4.821
#1	.6080	1.352	.4075	.3460	.4430
#2	.7032	1.179	.4115	.4407	.4148
#3	.6589	1.240	.3830	.3636	.3895
#4	.6429	1.202	.3665	.3536	.4275
#5	.7769	1.104	.4111	.4240	.4305

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4089	.3856
Stddev	.0345	.0437
%RSD	8.439	11.34
#1	.3881	.3719
#2	.4666	.4439
#3	.4155	.3408
#4	.3857	.3536
#5	.3884	.4176

Sample Name: s98t003389 Run Time: 03/03/99 15:47

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	103.06
Stddev	.38
%RSD	.37101

#1	103.08
#2	102.87
#3	102.69
#4	103.70
#5	102.98

Method: ACTINIDE Sample Name: s98t003389_d

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:51 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3601	.3497	4.838	193.6	.3730
Stdddev	.0440	.0383	.150	4.0	.0426
%RSD	12.21	10.94	3.108	2.082	11.43
#1	.4070	.3625	4.824	191.4	.3625
#2	.3925	.3614	4.875	198.3	.3676
#3	.3239	.2830	4.713	197.7	.3993
#4	.3720	.3812	4.703	190.8	.4242
#5	.3051	.3606	5.073	189.9	.3113

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.7001	1.088	.3799	.3692	.3777
Stdddev	.0452	.038	.0340	.0681	.0451
%RSD	6.450	3.464	8.939	18.46	11.94
#1	.7727	1.145	.3816	.4706	.4102
#2	.6884	1.084	.3271	.3022	.4050
#3	.6729	1.072	.4056	.3899	.3993
#4	.7101	1.042	.3720	.3720	.3013
#5	.6565	1.097	.4130	.3113	.3729

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3696	.3641
Stdddev	.0581	.0258
%RSD	15.73	7.091
#1	.4070	.3403
#2	.3302	.3520
#3	.4182	.3679
#4	.4058	.3535
#5	.2866	.4068

Sample Name: s98t003389_d Run Time: 03/03/99 15:51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	106.88
Stddev	1.50
%RSD	1.4076

#1	104.82
#2	107.00
#3	106.01
#4	108.43
#5	108.15

Method: ACTINIDE Sample Name: s98t003389_a

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 15:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3611	.5185	1551.	191.3	2.436
Stddev	.0181	.0372	24.	3.3	.086
%RSD	5.003	7.172	1.549	1.730	3.532

#1	.3569	.5135	1540.	193.3	2.423
#2	.3723	.5046	1527.	186.8	2.415
#3	.3713	.5740	1581.	193.6	2.524
#4	.3312	.4723	1534.	188.7	2.309
#5	.3741	.5281	1571.	194.0	2.508

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1968.	1783.	48.67	.6215	.3357
Stddev	36.	26.	1.29	.0233	.0283
%RSD	1.819	1.475	2.643	3.748	8.435

#1	1976.	1766.	46.76	.6481	.3444
#2	1918.	1763.	48.11	.6030	.2984
#3	2005.	1811.	50.05	.5928	.3151
#4	1947.	1761.	48.93	.6256	.3527
#5	1995.	1811.	49.49	.6381	.3678

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1962.	.3897
Stddev	25.	.0383
%RSD	1.268	9.825

#1	1950.	.3883
#2	1959.	.3384
#3	1991.	.3681
#4	1928.	.4263
#5	1981.	.4275

2742

Sample Name: s98t003389_a Run Time: 03/03/99 15:55

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.28
Stddev	1.17
%RSD	1.0945

#1	106.46
#2	108.35
#3	106.84
#4	108.69
#5	106.04

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 16:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3667	.3796	409.7	7.381	.9371
Stddev	.0261	.0209	8.8	.228	.0420
%RSD	7.129	5.511	2.151	3.093	4.486
#1	.3717	.4019	394.6	7.188	.9035
#2	.3280	.3659	410.6	7.743	.9462
#3	.4000	.3625	415.3	7.325	.8844
#4	.3739	.3647	411.5	7.447	.9684
#5	.3597	.4031	416.4	7.200	.9829

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	516.9	463.3	12.87	.4225	.3715
Stddev	10.5	13.0	.17	.0320	.0154
%RSD	2.038	2.797	1.288	7.571	4.141
#1	504.7	444.2	12.64	.4472	.3837
#2	525.4	473.3	12.80	.4447	.3532
#3	529.1	476.7	13.05	.4438	.3563
#4	508.4	458.1	12.83	.3984	.3831
#5	516.9	464.4	13.01	.3783	.3814

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	521.2	.3466
Stddev	4.7	.0268
%RSD	.9080	7.742
#1	515.4	.3082
#2	522.0	.3438
#3	528.3	.3781
#4	521.3	.3647
#5	519.2	.3380

2744

Sample Name: ccv Run Time: 03/03/99 16:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.79
Stddev	1.81
%RSD	1.6805
#1	110.32
#2	105.69
#3	106.66
#4	108.77
#5	107.50

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 16:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3396	.3810	.5065	.3978	.3423
Stddev	.0395	.0236	.0698	.0403	.0590
%RSD	11.63	6.182	13.77	10.13	17.24
#1	.3105	.4161	.5921	.4193	.3777
#2	.2996	.3889	.5706	.4112	.3570
#3	.3339	.3528	.4410	.3685	.2425
#4	.3989	.3704	.4749	.3451	.3419
#5	.3553	.3769	.4541	.4449	.3924

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3400	.6422	.3729	.3827	.3796
Stddev	.0579	.0313	.0261	.0182	.0369
%RSD	17.03	4.871	7.011	4.744	9.719
#1	.4097	.6369	.4129	.3937	.3713
#2	.3411	.5897	.3570	.3666	.3602
#3	.2488	.6646	.3654	.4032	.4441
#4	.3546	.6616	.3831	.3609	.3704
#5	.3460	.6581	.3460	.3893	.3522

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3676	.3889
Stddev	.0311	.0425
%RSD	8.459	10.93
#1	.4097	.3809
#2	.3347	.3347
#3	.3465	.4536
#4	.3577	.3831
#5	.3893	.3924

Sample Name: ccb Run Time: 03/03/99 16:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	105.55
Stddev	1.46
%RSD	1.3870

#1	104.14
#2	104.57
#3	105.82
#4	105.29
#5	107.89

2747

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 27714

Analyst: kjt

Instrument: ICPMS1

Book#: 11859

Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	@MSACT TH232	QC	1.0e-2	1.07e-02	107.000 % Recovery	
1	ICV	0	@MSACT NF237	QC	9.84e-3	1.17e-02	118.902 % Recovery	
1	ICV	0	@MSACT PU239	QC	1.00e-2	9.96e-03	99.600 % Recovery	
1	ICV	0	@MSACT 241	QC	1.03e-2	1.07e-02	103.883 % Recovery	
2	ICB	0	@MSAC-QC PA231	QC	1	<3.96e-6	ug/mL	
2	ICB	0	@MSAC-QC TH229	QC	1	<5.46e-6	ug/mL	
2	ICB	0	@MSAC-QC TH230	QC	1	<1.67e-6	ug/mL	
2	ICB	0	@MSAC-QC TH232	QC	1	<3.75e-6	ug/mL	
2	ICB	0	@MSAC-QC NF237	QC	1	<2.10e-6	ug/mL	
2	ICB	0	@MSAC-QC PU239	QC	1	<3.65e-6	ug/mL	
2	ICB	0	@MSAC-QC PU240	QC	1	<1.02e-6	ug/mL	
2	ICB	0	@MSAC-QC PU242	QC	1	<3.14e-6	ug/mL	
2	ICB	0	@MSAC-QC 241	QC	1	<1.68e-6	ug/mL	
2	ICB	0	@MSAC-QC 243	QC	1	<8.97e-7	ug/mL	
2	ICB	0	@MSAC-QC 244	QC	1	<5.74e-6	ug/mL	
3	BLNK-PREP	0	@MSACTA1 PA231-A1 SOLID		1	<4.36e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 TH229-A1 SOLID		1	<6.00e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 TH230-A1 SOLID		1	<1.84e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 TH232-A1 SOLID		1	<4.12e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 NP237-A1 SOLID		1	<2.31e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 PU239-A1 SOLID		1	<4.01e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 PU240-A1 SOLID		1	<1.12e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 PU242-A1 SOLID		1	<3.45e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 241-A1 SOLID		1	<1.85e-5	ug/g	
3	BLNK-PREP	0	@MSACTA1 243-A1 SOLID		1	<9.86e-6	ug/g	
3	BLNK-PREP	0	@MSACTA1 244-A1 SOLID		1	<6.31e-5	ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 PA231-A1 SOLID		N/A	< 3.151e-01	0.315 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 TH229-A1 SOLID		N/A	< 4.342e-01	0.434 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 TH230-A1 SOLID		N/A	< 1.330e-01	0.133 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 TH232-A1 SOLID		N/A	< 2.077e+00	0.298 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 NP237-A1 SOLID		N/A	< 1.669e-01	0.167 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 PU239-A1 SOLID		N/A	< 2.903e-01	0.290 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 PU240-A1 SOLID		N/A	< 8.086e-02	8.08e-002 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 PU242-A1 SOLID		N/A	< 2.498e-01	0.250 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 241-A1 SOLID		N/A	< 1.335e-02	0.134 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 243-A1 SOLID		N/A	< 7.134e-02	7.13e-002 ug/g	
4	SAMPLE	S98T002713	0 A @MSACTA1 244-A1 SOLID		N/A	< 4.564e-01	0.456 ug/g	
5	DUP	S98T002713	0 A @MSACTA1 PA231-A1 SOLID		<3.15e-1	<3.09e-1	RPD	
5	DUP	S98T002713	0 A @MSACTA1 TH229-A1 SOLID		<4.34e-1	<4.26e-1	RPD	
5	DUP	S98T002713	0 A @MSACTA1 TH230-A1 SOLID		<1.33e-1	<1.30e-1	RPD	
5	DUP	S98T002713	0 A @MSACTA1 TH232-A1 SOLID		2.08e+00	1.55e+00	29.201 RPD	

Units shown for QC (BLK/BKG) may not reflect the actual units.

2748

LABCORE Completed Worklist Report for Worklist# 27714

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 DUP	S98T002713	0 A	OMSACTA1 NP237-A1	SOLID	<1.67e-1	<1.64e-1		RPD
5 DUP	S98T002713	0 A	OMSACTA1 PU239-A1	SOLID	<2.90e-1	<2.85e-1		RPD
5 DUP	S98T002713	0 A	OMSACTA1 PU240-A1	SOLID	<8.09e-2	<7.93e-2		RPD
5 DUP	S98T002713	0 A	OMSACTA1 PU242-A1	SOLID	<2.50e-1	<2.45e-1		RPD
5 DUP	S98T002713	0 A	OMSACTA1 241-A1	SOLID	<1.34e-1	<1.31e-1		RPD
5 DUP	S98T002713	0 A	OMSACTA1 243-A1	SOLID	<7.13e-2	<7.00e-2		RPD
5 DUP	S98T002713	0 A	OMSACTA1 244-A1	SOLID	<4.56e-1	<4.47e-1		RPD
6 SAMPLE	S98T002846	0 A	OMSACTA1 PA231-A1	SOLID	N/A	< 1.623e-01	0.162	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 TH229-A1	SOLID	N/A	< 2.237e-01	0.224	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 TH230-A1	SOLID	N/A	< 6.849e-02	6.85e-002	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 TH232-A1	SOLID	N/A	< 2.529e+00	0.154	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 NP237-A1	SOLID	N/A	< 8.595e-02	8.59e-002	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 PU239-A1	SOLID	N/A	< 1.495e-01	0.149	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 PU240-A1	SOLID	N/A	< 4.165e-02	4.16e-002	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 PU242-A1	SOLID	N/A	< 1.287e-01	0.129	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 241-A1	SOLID	N/A	< 6.876e-02	6.88e-002	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 243-A1	SOLID	N/A	< 3.675e-02	3.67e-002	ug/g
6 SAMPLE	S98T002846	0 A	OMSACTA1 244-A1	SOLID	N/A	< 2.351e-01	0.235	ug/g
7 DUP	S98T002846	0 A	OMSACTA1 PA231-A1	SOLID	<1.62e-1	<1.61e-1		RPD
7 DUP	S98T002846	0 A	OMSACTA1 TH229-A1	SOLID	<2.24e-1	<2.21e-1		RPD
7 DUP	S98T002846	0 A	OMSACTA1 TH230-A1	SOLID	<6.85e-2	<6.78e-2		RPD
7 DUP	S98T002846	0 A	OMSACTA1 TH232-A1	SOLID	2.53e+00	2.72e+00	7.238	RPD
7 DUP	S98T002846	0 A	OMSACTA1 NP237-A1	SOLID	<8.59e-2	<8.51e-2		RPD
7 DUP	S98T002846	0 A	OMSACTA1 PU239-A1	SOLID	<1.49e-1	<1.48e-1		RPD
7 DUP	S98T002846	0 A	OMSACTA1 PU240-A1	SOLID	<4.16e-2	<4.12e-2		RPD
7 DUP	S98T002846	0 A	OMSACTA1 PU242-A1	SOLID	<1.29e-1	<1.27e-1		RPD
7 DUP	S98T002846	0 A	OMSACTA1 241-A1	SOLID	<6.88e-2	<6.81e-2		RPD
7 DUP	S98T002846	0 A	OMSACTA1 243-A1	SOLID	<3.67e-2	<3.64e-2		RPD
7 DUP	S98T002846	0 A	OMSACTA1 244-A1	SOLID	<2.35e-1	<2.33e-1		RPD
8 SPK-POST	S98T002846	0 A	OMSACTA1 TH232-A1	SOLID	0.020	1.99e-02	99.500	% Recovery
8 SPK-POST	S98T002846	0 A	OMSACTA1 NP237-A1	SOLID	0.01976	2.24e-02	113.360	% Recovery
8 SPK-POST	S98T002846	0 A	OMSACTA1 PU239-A1	SOLID	0.02008	1.96e-02	97.610	% Recovery
8 SPK-POST	S98T002846	0 A	OMSACTA1 241-A1	SOLID	0.02052	2.09e-02	101.852	% Recovery
9 SAMPLE	S98T002740	0 A	OMSACTA1 PA231-A1	SOLID	N/A	< 1.625e-01	0.163	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 TH229-A1	SOLID	N/A	< 2.240e-01	0.224	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 TH230-A1	SOLID	N/A	< 6.858e-02	6.86e-002	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 TH232-A1	SOLID	N/A	< 5.654e+00	0.154	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 NP237-A1	SOLID	N/A	< 8.606e-02	8.61e-002	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 PU239-A1	SOLID	N/A	< 1.497e-01	0.150	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 PU240-A1	SOLID	N/A	< 4.170e-02	4.17e-002	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 PU242-A1	SOLID	N/A	< 1.288e-01	0.129	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 241-A1	SOLID	N/A	< 6.884e-02	6.88e-002	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 243-A1	SOLID	N/A	< 3.680e-02	3.68e-002	ug/g
9 SAMPLE	S98T002740	0 A	OMSACTA1 244-A1	SOLID	N/A	< 2.354e-01	0.235	ug/g
10 DUP	S98T002740	0 A	OMSACTA1 PA231-A1	SOLID	<1.63e-1	<1.54e-1		RPD
10 DUP	S98T002740	0 A	OMSACTA1 TH229-A1	SOLID	<2.24e-1	<2.13e-1		RPD
10 DUP	S98T002740	0 A	OMSACTA1 TH230-A1	SOLID	<6.86e-2	<6.51e-2		RPD
10 DUP	S98T002740	0 A	OMSACTA1 TH232-A1	SOLID	5.65e+00	5.35e+00	5.455	RPD
10 DUP	S98T002740	0 A	OMSACTA1 NP237-A1	SOLID	<8.61e-2	<8.18e-2		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27714

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
10 DUP	S98T002740	0 A	@MSACT1 PU239-A1	SOLID	<1.50e-1	<1.42e-1		RPD
10 DUP	S98T002740	0 A	@MSACT1 PU240-A1	SOLID	<4.17e-2	<3.96e-2		RPD
10 DUP	S98T002740	0 A	@MSACT1 PU242-A1	SOLID	<1.29e-1	<1.22e-1		RPD
10 DUP	S98T002740	0 A	@MSACT1 241-A1	SOLID	<6.88e-2	<6.54e-2		RPD
10 DUP	S98T002740	0 A	@MSACT1 243-A1	SOLID	<3.68e-2	<3.50e-2		RPD
10 DUP	S98T002740	0 A	@MSACT1 244-A1	SOLID	<2.35e-1	<2.24e-1		RPD
11 CCV	0		@MSACT TH232	QC	1.0e-2	1.08e-02	108.000 % Recovery	
11 CCV	0		@MSACT NP237	QC	9.84e-3	1.17e-02	118.902 % Recovery	
11 CCV	0		@MSACT PU239	QC	1.00e-2	1.02e-02	102.000 % Recovery	
11 CCV	0		@MSACT 241	QC	1.03e-2	1.09e-02	105.825 % Recovery	
12 CCB	0		@MSAC-QC PA231	QC	1	<3.96e-6		ug/mL
12 CCB	0		@MSAC-QC TH239	QC	1	<5.46e-6		ug/mL
12 CCB	0		@MSAC-QC TH230	QC	1	<1.67e-6		ug/mL
12 CCB	0		@MSAC-QC TH232	QC	1	<3.75e-6		ug/mL
12 CCB	0		@MSAC-QC NP237	QC	1	<2.10e-6		ug/mL
12 CCB	0		@MSAC-QC PU239	QC	1	<3.65e-6		ug/mL
12 CCB	0		@MSAC-QC PU240	QC	1	<1.02e-6		ug/mL
12 CCB	0		@MSAC-QC PU242	QC	1	<3.14e-6		ug/mL
12 CCB	0		@MSAC-QC 241	QC	1	<1.68e-6		ug/mL
12 CCB	0		@MSAC-QC 243	QC	1	<8.97e-7		ug/mL
12 CCB	0		@MSAC-QC 244	QC	1	<5.74e-6		ug/mL

Final page for worklist# 27714

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

H. Anastro 3-4-99
Reviewer Signature _____ Date _____

Slightly high RPD on Th-232 (S98T002713)
is noted- will be rerun if requested

LABCORE Data Entry Template for Worklist# 27714

Analyst: K5 Instrument: ICPMS1 Book# 11855

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSACT	QC	← actually 11859 = not 11855	
2 ICB			@MSACTA1	SOLID		
3 BLNK-PREP			@MSACTA1	SOLID		
4 SAMPLE	S98T002713 0 A		@MSACTA1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
5 DUP	S98T002713 0 A		@MSACTA1	SOLID		
6 SAMPLE	S98T002846 0 A		@MSACTA1	SOLID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
7 DUP	S98T002846 0 A		@MSACTA1	SOLID		
8 SPK-POST	S98T002846 0 A		@MSACTA1	SOLID		
9 SAMPLE	S98T002740 0 A		@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
10 DUP	S98T002740 0 A		@MSACTA1	SOLID		
~~11 SPK-POST~~	~~S98T002740 0 A~~		~~@MSACTA1~~	~~SOLID~~		
12 CCV			@MSACT	QC		
13 CCB			@MSACTA1	SOLID		

Final page for worklist # 27714

K5 Romelle 03-02-99

Data Entry Comments:

LABCORE Data Entry Template for Worklist# 27714

S Type	Sample#	R A Test	Matrix	Group#	Project	Date
Analyst Signature		Date		Analyst Signature		
file	ac 0302a.txt					
	01LK-PREP	.1-10		1L df		
	898T002713	.05-10		201 df	Stream at	
	898T002713	.025-10		401 df	401 df	
	-d	.025-10		401 df		
	898T002840	.05-10		201 df		
	2840-d	.05-10		201 df		
	2840-a	.05-10		201 df +	.04 ml #1 8 BSS9	
	898T002740	.05-10		201 df		
	2740-d	.05-10		201 df		

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE

Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/02/99 13:15 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	190300.
Stddev	1238.
%RSD	.6505

#1	190700.
#2	192300.
#3	191300.
#4	188800.
#5	189300.
#6	190200.
#7	189300.
#8	191700.
#9	190400.
#10	188800.

48
3/2/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2153 TO 2167

Method: ACTINIDE Sample Name: cal blank

Operator: kjt

Comment: background

Run Time: 03/02/99 13:18 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2758	.3011	.3439	.3424	.2784
Stddev	.0605	.0188	.0434	.0140	.0452
%RSD	21.95	6.228	12.61	4.083	16.23

#1	.3280	.2811	.3568	.3424	.2523
#2	.2165	.3229	.4010	.3194	.2839
#3	.3493	.2858	.3387	.3564	.2188
#4	.2624	.2979	.2801	.3440	.3369
#5	.2226	.3180	.3427	.3498	.3003

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3083	.3679	.3140	.3077	.2999
Stddev	.0182	.0325	.0131	.0415	.0777
%RSD	5.907	8.840	4.180	13.48	25.92

#1	.3135	.3568	.3280	.3171	.3640
#2	.3194	.3904	.3052	.3478	.2058
#3	.3211	.3705	.3281	.2435	.3952
#4	.2766	.3191	.3085	.3369	.2660
#5	.3109	.4028	.3003	.2933	.2685

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3226	.3027
Stddev	.0219	.0120
%RSD	6.777	3.957

#1	.3460	.3171
#2	.3087	.2875
#3	.2929	.3105
#4	.3262	.2943
#5	.3392	.3039

Sample Name: cal blank Run Time: 03/02/99 13:18

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	93.846
Stddev	.790
%RSD	.84207

#1	92.491
#2	93.928
#3	94.471
#4	94.002
#5	94.337

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.010 ppm cal

Run Time: 03/02/99 13:22 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1878	.2732	810.8	879.7	1.105
Stddev	.0193	.0105	15.4	19.5	.052
%RSD	10.25	3.827	1.901	2.220	4.700

#1	.2092	.2622	797.9	858.3	1.105
#2	.2060	.2652	822.7	871.8	1.064
#3	.1645	.2763	801.6	896.0	1.180
#4	.1762	.2735	800.1	868.1	1.127
#5	.1832	.2888	831.9	904.3	1.050

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.326	.5177	.1897	.1846	.2108
Stddev	.058	.0430	.0175	.0448	.0201
%RSD	4.393	8.306	9.241	24.27	9.559

#1	1.299	.5682	.2061	.2404	.2217
#2	1.414	.5305	.1747	.1654	.2403
#3	1.350	.5153	.2018	.1211	.1925
#4	1.261	.4497	.1671	.2066	.1945
#5	1.308	.5249	.1988	.1895	.2050

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2169	.2021
Stddev	.0196	.0181
%RSD	9.035	8.936

#1	.2435	.2029
#2	.2278	.2309
#3	.1925	.1832
#4	.2066	.1914
#5	.2143	.2019

Sample Name: cal std Run Time: 03/02/99 13:22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.60
Stddev	1.21
%RSD	1.1237
#1	106.77
#2	106.82
#3	107.38
#4	109.70
#5	107.33

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.100 ppm

Run Time: 03/02/99 13:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2660	.9160	8125.	8507.	9.713
Stddev	.0198	.0553	149.	222.	.269
%RSD	7.451	6.033	1.838	2.611	2.771
#1	.2508	.8611	7907.	8184.	9.575
#2	.2436	.8808	8301.	8677.	10.19
#3	.2819	1.005	8083.	8415.	9.608
#4	.2634	.9159	8114.	8514.	9.558
#5	.2900	.9169	8217.	8744.	9.633

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	11.52	2.111	.1939	.2057	.2025
Stddev	.23	.066	.0090	.0368	.0390
%RSD	2.036	3.112	4.656	17.91	19.24
#1	11.16	2.190	.1934	.2145	.1752
#2	11.68	2.168	.1936	.1999	.2030
#3	11.49	2.093	.2084	.1502	.2697
#4	11.49	2.068	.1838	.2114	.1807
#5	11.77	2.036	.1902	.2526	.1840

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1730	.1837
Stddev	.0196	.0332
%RSD	11.34	18.09
#1	.1692	.1541
#2	.2030	.2343
#3	.1778	.1869
#4	.1654	.1532
#5	.1497	.1902

Sample Name: cal std Run Time: 03/02/99 13:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	108.30
Stddev	1.51
%RSD	1.3909
#1	110.33
#2	106.72
#3	108.77
#4	108.81
#5	106.89

Method: ACTINIDE Sample Name: icv

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 13:45 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2208	.3174	865.5	13.90	1.331
Stddev	.0370	.0458	25.2	.49	.106
%RSD	16.78	14.42	2.913	3.501	7.952

#1	.2240	.2986	850.1	13.21	1.324
#2	.1585	.2708	873.7	13.85	1.176
#3	.2431	.3252	903.1	14.10	1.366
#4	.2243	.3917	863.7	14.55	1.319
#5	.2539	.3007	836.8	13.78	1.470

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1018.	888.0	24.24	.3725	.2760
Stddev	31.	17.6	.54	.0213	.0320
%RSD	3.035	1.977	2.227	5.714	11.58

#1	989.0	867.4	23.96	.3538	.2305
#2	1024.	873.9	24.19	.3963	.2741
#3	1065.	909.6	24.52	.3766	.2636
#4	1019.	899.8	24.97	.3883	.3113
#5	991.8	889.3	23.55	.3475	.3007

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	973.1	.2543
Stddev	17.7	.0387
%RSD	1.822	15.21

#1	959.4	.1883
#2	965.1	.2807
#3	1004.	.2739
#4	969.8	.2511
#5	967.1	.2773

Sample Name: icv Run Time: 03/02/99 13:45

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	100.07
Stddev	1.95
%RSD	1.9494

#1	102.69
#2	100.92
#3	97.371
#4	99.571
#5	99.772

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ich

Operator: kjt

Comment: 2% hno3

Run Time: 03/02/99 14:00 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2384	.2428	.4017	.3112	.2299
Stddev	.0445	.0360	.0454	.0270	.0272
%RSD	18.66	14.84	11.30	8.674	11.82

#1	.1860	.1954	.4160	.3246	.1923
#2	.2056	.2956	.4337	.3437	.2120
#3	.2784	.2467	.3796	.2815	.2372
#4	.2339	.2308	.3337	.3212	.2557
#5	.2882	.2456	.4453	.2849	.2521

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2551	.3632	.2491	.2681	.2508
Stddev	.0415	.0572	.0201	.0353	.0257
%RSD	16.28	15.75	8.056	13.18	10.24

#1	.2458	.3814	.2175	.2490	.2711
#2	.2249	.4401	.2699	.2185	.2088
#3	.2784	.3638	.2594	.3037	.2436
#4	.2121	.3493	.2433	.2713	.2651
#5	.3143	.2816	.2554	.2980	.2652

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2716	.2691
Stddev	.0374	.0315
%RSD	13.77	11.72

#1	.2900	.3183
#2	.2699	.2570
#3	.2341	.2341
#4	.3244	.2776
#5	.2390	.2587

Sample Name: lcb Run Time: 03/02/99 14:00

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.72
Stddev	1.98
%RSD	1.8889

#1	105.76
#2	103.76
#3	105.38
#4	106.88
#5	101.80

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: blink-prep

Operator: kjt

Comment: 11 df

Run Time: 03/02/99 14:04 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2477	.2736	.4762	.7034	.2125
Stddev	.0287	.0190	.0716	.0591	.0127
%RSD	11.57	6.951	15.03	8.396	5.983

#1	.2789	.2789	.5976	.7389	.2101
#2	.2623	.2551	.4773	.7834	.2077
#3	.2630	.2910	.4173	.6838	.1964
#4	.2164	.2909	.4329	.6812	.2306
#5	.2179	.2519	.4562	.6298	.2179

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2238	.4427	.2636	.2529	.2591
Stddev	.0264	.0226	.0205	.0252	.0591
%RSD	11.78	5.094	7.759	9.980	22.81

#1	.1883	.4382	.2535	.2316	.2427
#2	.2296	.4482	.2878	.2368	.2040
#3	.2560	.4067	.2700	.2665	.3541
#4	.2377	.4541	.2342	.2909	.2732
#5	.2077	.4664	.2723	.2383	.2213

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2520	.2585
Stddev	.0249	.0161
%RSD	9.863	6.211

#1	.2499	.2680
#2	.2769	.2623
#3	.2630	.2525
#4	.2590	.2342
#5	.2111	.2757

HNF-1663 REV. 0

Sample Name: blink-prep Run Time: 03/02/99 14:04

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	94.089
Stddev	2.580
%RSD	2.7419

#1	92.035
#2	91.483
#3	95.063
#4	93.950
#5	97.916

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: s98t002713

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1949	.2067	4.574	8454.	.2055
Stddev	.0181	.0151	.080	89.	.0361
%RSD	9.305	7.284	1.750	1.050	17.58

#1	.2017	.1866	4.578	8484.	.1956
#2	.1767	.2277	4.679	8371.	.1468
#3	.1743	.2068	4.488	8355.	.2216
#4	.2128	.2006	4.505	8563.	.2280
#5	.2089	.2119	4.623	8497.	.2358

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	11.79	2.578	.2404	.2229	.2212
Stddev	.24	.063	.0308	.0152	.0226
%RSD	2.000	2.433	12.83	6.829	10.20

#1	11.96	2.576	.2950	.2107	.2318
#2	11.65	2.561	.2217	.2187	.2396
#3	11.66	2.609	.2334	.2452	.1832
#4	12.11	2.486	.2280	.2310	.2189
#5	11.55	2.656	.2238	.2089	.2328

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2265	.2306
Stddev	.0416	.0187
%RSD	18.36	8.088

#1	.1625	.2528
#2	.2067	.2247
#3	.2541	.2423
#4	.2614	.2037
#5	.2477	.2298

Sample Name: s98t002713 Run Time: 03/02/99 14:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.24
Stddev	1.17
%RSD	1.0481

#1	110.75
#2	111.28
#3	112.82
#4	109.66
#5	111.69

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002713_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:13 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2370	.2497	1.948	4116.	.2601
Stddev	.0359	.0776	1.677	5608.	.0465
%RSD	15.15	31.07	86.09	136.2	17.88

#1	.1967	.2088	3.727	10420.	.1967
#2	.2346	.1693	3.822	10100.	.2405
#3	.2449	.3649	1.028	60.89	.3121
#4	.2165	.2165	.6070	1.90	.2526
#5	.2921	.2889	.5583	1.25	.2986

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.871	1.665	.2653	.2553	.2833
Stddev	7.557	1.228	.0967	.1182	.0620
%RSD	128.7	73.73	36.44	46.30	21.89

#1	14.49	3.165	.2481	.1725	.2299
#2	13.80	2.821	.1485	.1188	.2524
#3	.5234	1.028	.4178	.4176	.3890
#4	.2789	.6792	.2526	.2494	.2822
#5	.2629	.6330	.2597	.3181	.2629

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3003	.2489
Stddev	.1129	.1035
%RSD	37.59	41.59

#1	.2723	.2118
#2	.1277	.1039
#3	.4226	.3649
#4	.3740	.2330
#5	.3051	.3311

HNF-1663 REV. 0

Sample Name: s98t002713_d Run Time: 03/02/99 14:13

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	99.228
Stddev	17.291
%RSD	17.426

#1	110.18
#2	112.26
#3	69.416
#4	101.60
#5	102.69

Method: ACTINIDE Sample Name: s98t002713

Operator: kjt

Comment: 401 df

Run Time: 03/02/99 14:22 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1957	.2182	2.460	4419.	.2194
Stddev	.0168	.0207	.061	7.	.0200
%RSD	8.573	9.477	2.496	.1566	9.097

#1	.1876	.2027	2.554	4422.	.1997
#2	.1970	.1909	2.388	4420.	.2061
#3	.1743	.2293	2.464	4426.	.2262
#4	.2199	.2412	2.428	4419.	.2504
#5	.1996	.2268	2.468	4408.	.2147

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5.992	1.421	.2376	.2479	.2164
Stddev	.051	.089	.0175	.0454	.0299
%RSD	.8423	6.249	7.353	18.31	13.80

#1	5.949	1.362	.2511	.2269	.1664
#2	5.946	1.431	.2364	.2031	.2455
#3	6.002	1.305	.2079	.2415	.2293
#4	5.994	1.478	.2443	.2443	.2168
#5	6.070	1.527	.2480	.3236	.2238

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2285	.2491
Stddev	.0200	.0262
%RSD	8.743	10.51

#1	.2239	.2451
#2	.2425	.2546
#3	.2538	.2110
#4	.2199	.2504
#5	.2027	.2843

Sample Name: s98t002713 Run Time: 03/02/99 14:22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.71
Stddev	.57
%RSD	.51770

#1	110.16
#2	109.98
#3	109.03
#4	109.16
#5	110.21

Method: ACTINIDE Sample Name: s98t002713_d

Operator: kjt

Comment: 401 df

Run Time: 03/02/99 14:25 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2199	.2201	1.955	5137.	.2207
Stddev	.0288	.0318	.109	32.	.0229
%RSD	13.09	14.43	5.581	.6319	10.38

#1	.1869	.2114	1.982	5161.	.2237
#2	.2030	.1878	1.936	5100.	.2151
#3	.2109	.2720	1.785	5164.	.2261
#4	.2570	.2238	1.990	5102.	.2510
#5	.2418	.2055	2.082	5155.	.1874

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7.175	1.639	.2328	.2382	.2140
Stddev	.178	.055	.0175	.0184	.0517
%RSD	2.476	3.355	7.499	7.731	24.15

#1	7.415	1.621	.2145	.2298	.2114
#2	7.018	1.630	.2363	.2242	.2120
#3	7.166	1.736	.2597	.2414	.2414
#4	7.282	1.600	.2207	.2268	.1331
#5	6.994	1.611	.2327	.2690	.2720

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2565	.2365
Stddev	.0213	.0228
%RSD	8.301	9.655

#1	.2696	.2727
#2	.2363	.2302
#3	.2475	.2414
#4	.2419	.2117
#5	.2871	.2267

Sample Name: s98t002713_d Run Time: 03/02/99 14:25

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.69
Stddev	.70
%RSD	.63690

#1	108.79
#2	110.04
#3	109.08
#4	110.23
#5	110.28

Method: ACTINIDE Sample Name: s98t002846

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:29 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2154	.2258	5.345	5132.	.2115
Stddev	.0195	.0288	.333	204.	.0293
%RSD	9.038	12.74	6.225	3.986	13.86

#1	.2064	.2202	4.758	4769.	.1844
#2	.2259	.2500	5.515	5192.	.2590
#3	.2421	.2421	5.525	5255.	.1937
#4	.2115	.1782	5.399	5235.	.2024
#5	.1910	.2387	5.529	5210.	.2178

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6.948	1.823	.2426	.2223	.2264
Stddev	.146	.140	.0371	.0260	.0394
%RSD	2.102	7.684	15.30	11.68	17.39

#1	6.704	1.665	.2559	.2147	.2339
#2	7.064	1.702	.1807	.1868	.2229
#3	7.057	1.888	.2451	.2179	.2572
#4	6.970	2.006	.2508	.2357	.2568
#5	6.947	1.856	.2805	.2566	.1611

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2210	.2146
Stddev	.0140	.0508
%RSD	6.332	23.68

#1	.2257	.1871
#2	.2349	.1837
#3	.2300	.3026
#4	.2145	.2145
#5	.1999	.1850

Sample Name: s98t002846 Run Time: 03/02/99 14:29

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	112.80
Stddev	4.69
%RSD	4.1622

#1	121.13
#2	110.66
#3	110.16
#4	110.33
#5	111.71

Method: ACTINIDE Sample Name: s98t002846_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:33 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2145	.1935	5.781	5292.	.2042
Stddev	.0257	.0207	.126	99.	.0236
%RSD	11.97	10.70	2.175	1.862	11.55

#1	.1696	.1725	5.722	5138.	.1871
#2	.2168	.1867	5.824	5373.	.2198
#3	.2277	.1828	5.729	5282.	.1768
#4	.2318	.1991	5.652	5285.	.2347
#5	.2264	.2264	5.977	5383.	.2023

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7.308	1.861	.2663	.2070	.2290
Stddev	.134	.074	.0453	.0254	.0278
%RSD	1.830	3.972	17.02	12.25	12.12

#1	7.222	1.816	.2368	.2427	.2456
#2	7.399	1.858	.3011	.2198	.2138
#3	7.195	1.789	.3266	.1798	.1888
#4	7.226	1.982	.2466	.1872	.2585
#5	7.499	1.863	.2204	.2053	.2385

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2281	.2297
Stddev	.0391	.0218
%RSD	17.13	9.496

#1	.2193	.2456
#2	.2740	.2289
#3	.1768	.2547
#4	.2110	.1991
#5	.2596	.2204

Analysis Report

03/02/99 14:37:12

page 2

HNF-1663 REV. 0

Sample Name: s98t002846_d Run Time: 03/02/99 14:33

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.71
Stddev	1.45
%RSD	1.2983

#1	114.00
#2	110.69
#3	111.24
#4	112.18
#5	110.42

Method: ACTINIDE Sample Name: s98t002846_a

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:40 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2349	.3632	1622.	5366.	2.183
Stddev	.0241	.0106	14.	55.	.058
%RSD	10.26	2.931	.8576	1.023	2.663

#1	.2201	.3588	1628.	5322.	2.216
#2	.2088	.3481	1621.	5345.	2.246
#3	.2243	.3637	1615.	5317.	2.170
#4	.2600	.3688	1605.	5442.	2.092
#5	.2610	.3764	1642.	5404.	2.188

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1954.	1743.	46.99	.4388	.2476
Stddev	7.	27.	.71	.0627	.0150
%RSD	.3445	1.555	1.509	14.30	6.059

#1	1945.	1707.	45.94	.3980	.2382
#2	1961.	1735.	47.17	.3814	.2361
#3	1959.	1732.	47.18	.4789	.2425
#4	1958.	1768.	47.89	.5291	.2479
#5	1950.	1773.	46.80	.4067	.2732

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1908.	.2639
Stddev	34.	.0291
%RSD	1.768	11.01

#1	1874.	.2684
#2	1898.	.2663
#3	1916.	.2334
#4	1890.	.3084
#5	1962.	.2428

HNF-1663 REV. 0

Sample Name: s98t002846_a Run Time: 03/02/99 14:40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	110.15
Stddev	.28
%RSD	.25001

#1	110.54
#2	110.13
#3	109.97
#4	110.26
#5	109.83

Method: ACTINIDE Sample Name: s98t002740

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:54 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2161	.2212	11.51	442.5	.2329
Stddev	.0282	.0142	.22	3.9	.0264
%RSD	13.03	6.403	1.889	.8845	11.34

#1	.2379	.1997	11.87	442.9	.2614
#2	.1813	.2310	11.37	446.6	.2339
#3	.1906	.2362	11.40	437.9	.2020
#4	.2424	.2167	11.35	439.2	.2110
#5	.2281	.2225	11.57	446.0	.2563

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8652	.7219	.2578	.2159	.2191
Stddev	.0692	.0647	.0379	.0317	.0281
%RSD	8.002	8.965	14.70	14.68	12.81

#1	.8606	.8107	.2614	.2467	.2203
#2	.8537	.6929	.2280	.2222	.2193
#3	.8337	.6858	.2703	.1963	.2163
#4	.9809	.7670	.2167	.2424	.2595
#5	.7970	.6534	.3126	.1718	.1802

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2131	.2377
Stddev	.0189	.0309
%RSD	8.862	13.01

#1	.1880	.2408
#2	.2134	.2894
#3	.2362	.2191
#4	.2024	.2281
#5	.2253	.2112

Sample Name: s98t002740 Run Time: 03/02/99 14:54

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	115.98
Stddev	2.12
%RSD	1.8254

#1	113.49
#2	114.02
#3	117.14
#4	116.90
#5	118.36

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: s98t002740_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 14:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2061	.2068	11.47	427.2	.2207
Stddev	.0159	.0191	.27	4.4	.0433
%RSD	7.722	9.257	2.382	1.026	19.64

#1	.2031	.2289	11.22	425.2	.1602
#2	.2222	.1941	11.41	422.0	.2700
#3	.1918	.2143	11.20	428.8	.1946
#4	.2232	.1808	11.76	426.4	.2345
#5	.1903	.2159	11.75	433.8	.2443

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8281	.6535	.2089	.2152	.1965
Stddev	.0278	.0372	.0160	.0092	.0314
%RSD	3.360	5.689	7.651	4.263	15.97

#1	.8440	.6352	.1917	.2060	.1574
#2	.8325	.7144	.2025	.2194	.1856
#3	.8009	.6204	.2284	.2228	.1974
#4	.7995	.6356	.2232	.2232	.1977
#5	.8636	.6619	.1989	.2045	.2443

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2134	.2278
Stddev	.0282	.0308
%RSD	13.21	13.51

#1	.1803	.2518
#2	.1941	.1941
#3	.2538	.2425
#4	.2203	.1949
#5	.2187	.2557

Sample Name: s98t002740_d Run Time: 03/02/99 14:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	117.71
Stddev	.80
%RSD	.68262

#1	116.50
#2	118.52
#3	118.21
#4	118.00
#5	117.34

Method: ACTINIDE Sample Name: ccv

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 15:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2237	.3163	871.4	13.87	1.222
Stddev	.0262	.0203	8.0	.17	.045
%RSD	11.70	6.424	.9168	1.229	3.693
#1	.2247	.3111	868.4	13.99	1.210
#2	.2263	.3308	873.0	13.72	1.163
#3	.2224	.3306	862.5	14.09	1.270
#4	.2594	.2829	869.2	13.84	1.264
#5	.1855	.3262	884.0	13.70	1.200

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1013.	909.9	24.39	.3181	.2175
Stddev	11.	16.4	.43	.0344	.0271
%RSD	1.087	1.806	1.778	10.82	12.45
#1	1013.	894.8	24.17	.2823	.2189
#2	1015.	901.9	23.86	.3714	.2031
#3	1008.	904.0	24.43	.3072	.2077
#4	998.8	911.7	24.44	.3301	.1945
#5	1029.	937.3	25.03	.2993	.2633

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	997.1	.2326
Stddev	18.8	.0168
%RSD	1.888	7.215
#1	993.5	.2506
#2	992.8	.2089
#3	986.4	.2341
#4	983.1	.2240
#5	1030.	.2454

Sample Name: ccv Run Time: 03/02/99 15:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	113.80
Stddev	1.67
%RSD	1.4679

#1	115.71
#2	114.89
#3	113.92
#4	113.10
#5	111.39

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccb

Operator: kjt

Comment: 2% hno3

Run Time: 03/02/99 15:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2260	.2412	.3637	.3267	.2067
Stddev	.0220	.0405	.0245	.0315	.0127
%RSD	9.721	16.78	6.727	9.632	6.125

#1	.1998	.1851	.3732	.3467	.2174
#2	.2176	.2992	.3566	.2961	.1994
#3	.2195	.2439	.3781	.2896	.2104
#4	.2348	.2348	.3244	.3430	.1885
#5	.2585	.2429	.3862	.3581	.2180

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2460	.3517	.2417	.2458	.2453
Stddev	.0340	.0384	.0223	.0279	.0137
%RSD	13.83	10.91	9.213	11.36	5.593

#1	.2204	.3556	.2145	.2263	.2380
#2	.2720	.2961	.2327	.2780	.2266
#3	.2195	.3903	.2713	.2378	.2470
#4	.2256	.3337	.2565	.2719	.2626
#5	.2927	.3830	.2336	.2149	.2523

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2369	.2277
Stddev	.0395	.0483
%RSD	16.66	21.22

#1	.2380	.1969
#2	.2810	.2025
#3	.2653	.3110
#4	.2194	.2286
#5	.1806	.1993

Sample Name: ccb Run Time: 03/02/99 15:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.60
Stddev	2.49
%RSD	2.2733

#1	113.44
#2	110.31
#3	109.33
#4	107.88
#5	107.04

LABCORE Completed Worklist Report for Worklist# 27715

Analyst: bjg

Instrument: ICPMS1

Book#: 11B59

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	0	QMSACT TH232	QC	1.00e-2	1.07e-02	107.000 % Recovery	
1 ICV	0	0	QMSACT NP237	QC	9.88e-3	1.17e-02	118.421 % Recovery	
1 ICV	0	0	QMSACT PU239	QC	1.004e-2	1.01e-02	100.598 % Recovery	
1 ICV	0	0	QMSACT 241	QC	1.026e-2	1.07e-02	104.288 % Recovery	
2 ICB	0	0	QMSACT1 TH229-A1	SOLID	1	<2.30e-6	ug/mL	
2 ICB	0	0	QMSACT1 TH230-A1	SOLID	1	<1.94e-6	ug/mL	
2 ICB	0	0	QMSACT1 TH232-A1	SOLID	1	<2.29e-6	ug/mL	
3 BLNK-PREP	0	0	QMSACT1 PA231-A1	SOLID	1	<4.06e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 TH229-A1	SOLID	1	<2.52e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 TH230-A1	SOLID	1	<2.13e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 TH232-A1	SOLID	1	2.77e-05	2.770e-005 ug/g	
3 BLNK-PREP	0	0	QMSACT1 NP237-A1	SOLID	1	<2.60e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 PU239-A1	SOLID	1	<3.64e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 PU240-A1	SOLID	1	<2.83e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 PU242-A1	SOLID	1	<1.81e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 241-A1	SOLID	1	<2.74e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 243-A1	SOLID	1	<1.18e-5	ug/g	
3 BLNK-PREP	0	0	QMSACT1 244-A1	SOLID	1	<2.04e-5	ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 PA231-A1	SOLID	N/A	< 1.513e-01	0.151 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 TH229-A1	SOLID	N/A	< 9.407e-02	9.41e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 TH230-A1	SOLID	N/A	< 7.942e-02	7.94e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 TH232-A1	SOLID	N/A	< 5.281e-01	9.37e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 NP237-A1	SOLID	N/A	< 9.696e-02	9.70e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 PU239-A1	SOLID	N/A	< 1.358e-01	0.136 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 PU240-A1	SOLID	N/A	< 1.055e-01	0.105 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 PU242-A1	SOLID	N/A	< 6.743e-02	6.74e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 241-A1	SOLID	N/A	< 1.023e-01	0.102 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 243-A1	SOLID	N/A	< 4.398e-02	4.40e-002 ug/g	
4 SAMPLE	S98T002928	0 A	QMSACT1 244-A1	SOLID	N/A	< 7.589e-02	7.59e-002 ug/g	
5 DUP	S98T002928	0 A	QMSACT1 PA231-A1	SOLID	<1.51e-1	<1.41e-1	RPD	
5 DUP	S98T002928	0 A	QMSACT1 TH229-A1	SOLID	<9.41e-2	<8.76e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 TH230-A1	SOLID	<7.94e-2	<7.40e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 TH232-A1	SOLID	5.28e-01	4.59e-01	13.982 RPD	
5 DUP	S98T002928	0 A	QMSACT1 NP237-A1	SOLID	<9.70e-2	<9.03e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 PU239-A1	SOLID	<1.36e-1	<1.27e-1	RPD	
5 DUP	S98T002928	0 A	QMSACT1 PU240-A1	SOLID	<1.05e-1	<9.83e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 PU242-A1	SOLID	<6.74e-2	<6.28e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 241-A1	SOLID	<1.02e-1	<9.53e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 243-A1	SOLID	<4.40e-2	<4.10e-2	RPD	
5 DUP	S98T002928	0 A	QMSACT1 244-A1	SOLID	<7.59e-2	<7.07e-2	RPD	
6 SAMPLE	S98T002929	0 A	QMSACT1 PA231-A1	SOLID	N/A	< 1.459e-01	0.146 ug/g	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27715

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002929	0 A	OMSACTA1 TH229-A1	SOLID	N/A <	9.074e-02	9.07e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 TH230-A1	SOLID	N/A <	7.660e-02	7.66e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 TH232-A1	SOLID	N/A	6.305e-01	9.03e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 NP237-A1	SOLID	N/A <	9.353e-02	9.35e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 PU239-A1	SOLID	N/A <	1.310e-01	0.131	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 PU240-A1	SOLID	N/A <	1.018e-01	0.102	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 PU242-A1	SOLID	N/A <	6.504e-02	6.50e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 241-A1	SOLID	N/A <	9.865e-02	9.87e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 243-A1	SOLID	N/A <	4.243e-02	4.24e-002	ug/g
6 SAMPLE	S98T002929	0 A	OMSACTA1 244-A1	SOLID	N/A <	7.320e-02	7.32e-002	ug/g
7 DUP	S98T002929	0 A	OMSACTA1 PA231-A1	SOLID	<1.46e-1	<1.51e-1		RPD
7 DUP	S98T002929	0 A	OMSACTA1 TH229-A1	SOLID	<9.07e-2	<9.38e-2		RPD
7 DUP	S98T002929	0 A	OMSACTA1 TH230-A1	SOLID	<7.66e-2	<7.92e-2		RPD
7 DUP	S98T002929	0 A	OMSACTA1 TH232-A1	SOLID	6.30e-01	6.95e-01	9.811	RPD
7 DUP	S98T002929	0 A	OMSACTA1 NP237-A1	SOLID	<9.35e-2	<9.66e-2		RPD
7 DUP	S98T002929	0 A	OMSACTA1 PU239-A1	SOLID	<1.31e-1	<1.35e-1		RPD
7 DUP	S98T002929	0 A	OMSACTA1 PU240-A1	SOLID	<1.02e-1	<1.05e-1		RPD
7 DUP	S98T002929	0 A	OMSACTA1 PU242-A1	SOLID	<6.50e-2	<6.72e-2		RPD
7 DUP	S98T002929	0 A	OMSACTA1 241-A1	SOLID	<9.87e-2	<1.02e-1		RPD
7 DUP	S98T002929	0 A	OMSACTA1 243-A1	SOLID	<4.24e-2	<4.38e-2		RPD
7 DUP	S98T002929	0 A	OMSACTA1 244-A1	SOLID	<7.32e-2	<7.56e-2		RPD
8 SAMPLE	S98T002930	0 A	OMSACTA1 PA231-A1	SOLID	N/A <	5.912e+00	5.912	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 TH229-A1	SOLID	N/A <	3.677e+00	3.677	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 TH230-A1	SOLID	N/A <	3.104e+00	3.104	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 TH232-A1	SOLID	N/A <	3.661e+00	3.661	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 NP237-A1	SOLID	N/A <	3.790e+00	3.790	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 PU239-A1	SOLID	N/A <	5.308e+00	5.308	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 PU240-A1	SOLID	N/A <	4.124e+00	4.124	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 PU242-A1	SOLID	N/A <	2.636e+00	2.635	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 241-A1	SOLID	N/A <	3.997e+00	3.998	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 243-A1	SOLID	N/A <	1.719e+00	1.719	ug/g
8 SAMPLE	S98T002930	0 A	OMSACTA1 244-A1	SOLID	N/A <	2.966e+00	2.966	ug/g
9 DUP	S98T002930	0 A	OMSACTA1 PA231-A1	SOLID	<5.91e0	<6.01e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 TH229-A1	SOLID	<3.68e0	<3.74e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 TH230-A1	SOLID	<3.10e0	<3.15e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 TH232-A1	SOLID	<3.66e0	<3.72e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 NP237-A1	SOLID	<3.79e0	<3.85e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 PU239-A1	SOLID	<5.31e0	<5.39e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 PU240-A1	SOLID	<4.12e0	<4.19e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 PU242-A1	SOLID	<2.64e0	<2.68e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 241-A1	SOLID	<4.00e0	<4.06e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 243-A1	SOLID	<1.72e0	<1.75e0		RPD
9 DUP	S98T002930	0 A	OMSACTA1 244-A1	SOLID	<2.97e0	<3.01e0		RPD
10 SPK-POST	S98T002930	0 A	OMSACTA1 TH232-A1	SOLID	0.020	1.97e-02	98.500	% Recovery
10 SPK-POST	S98T002930	0 A	OMSACTA1 NP237-A1	SOLID	0.01976	2.22e-02	112.348	% Recovery
10 SPK-POST	S98T002930	0 A	OMSACTA1 PU239-A1	SOLID	0.02008	1.93e-02	96.116	% Recovery
10 SPK-POST	S98T002930	0 A	OMSACTA1 241-A1	SOLID	0.02052	2.07e-02	100.877	% Recovery
11 CCV		0	OMSACT TH232	QC	1.00e-2	1.05e-02	105.000	% Recovery
11 CCV		0	OMSACT NP237	QC	9.88e-3	1.19e-02	120.445	% Recovery

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27715

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	CCV	0	@MSACT	P0239	QC		1.004e-2	1.04e-02	103.586 % Recovery	
11	CCV	0	@MSACT	241	QC		1.026e-2	1.10e-02	107.212 % Recovery	
12	CCB	0	@MSACTA1	PA231-A1	LIQUID		1	<3.69e-6		ug/mL
12	CCB	0	@MSACTA1	TH229-A1	LIQUID		1	<2.30e-6		ug/mL
12	CCB	0	@MSACTA1	TH230-A1	LIQUID		1	<1.94e-6		ug/mL
12	CCB	0	@MSACTA1	TH232-A1	LIQUID		1	<2.29e-6		ug/mL
12	CCB	0	@MSACTA1	P0239-A1	LIQUID		1	<3.31e-6		ug/mL
12	CCB	0	@MSACTA1	P0240-A1	LIQUID		1	<2.57e-6		ug/mL
12	CCB	0	@MSACTA1	P0242-A1	LIQUID		1	<1.65e-6		ug/mL
12	CCB	0	@MSACTA1	241-A1	LIQUID		1	<2.50e-6		ug/mL
12	CCB	0	@MSACTA1	243-A1	LIQUID		1	<1.07e-6		ug/mL
12	CCB	0	@MSACTA1	244-A1	LIQUID		1	<1.85e-6		ug/mL

Final page for worklist# 27715

Analyst Signature

Date

Analyst Signature

Date

03/04/99

Reviewer Signature

Date

12/29/98 13:34
A-0004-I

file ac030ae.txt

Page: 1

LABCORE Data Entry Template for Worklist# 27715

Analyst: KST Instrument: ICPMS1 Book# 11859Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSACT	QC		
2	ICB		@MSACTA1	SOLID		
3	BLNK-PREP		@MSACTA1	SOLID		
4	SAMPLE	S98T002928 0 A	@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
5	DUP	S98T002928 0 A	@MSACTA1	SOLID		
6	SPK-POST	~~S98T002928 0 A~~	~~@MSACTA1~~	~~SOLID~~		
7	SAMPLE	S98T002929 0 A	@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
8	DUP	S98T002929 0 A	@MSACTA1	SOLID		
9	SAMPLE	S98T002930 0 A	@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
10	DUP	S98T002930 0 A	@MSACTA1	SOLID		
11	SPK-POST	S98T002930 0 A	@MSACTA1	SOLID		
12	CCV		@MSACT	QC		
13	CCB		@MSACTA1	SOLID		

KST 03-02-99

Final page for worklist # 27715

K5 Piondeke 03-02-99

Data Entry Comments:

Uploaded & Validated by:

Saul R. Pang

03/04/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

12/29/98 13:34

A-0004-1

LABCORE Data Entry Template for Worklist# 27715

S Type	Sample#	R A Test	Matrix	Group#	Project	Date
Analyst Signature		Date		Analyst Signature		

File ac0302c.txt

Blnk -Prep   1-10   11 df
 S98T002928   .050-10   201 df
     2928-d   ↓   ↓
     2929   ↓   ↓
     2929-d   ↓   ↓
     2930   .05-10-250-10   8241 df
     2930-d   ↓   ↓
     2930-a   ↓   ↓
                  .04 ml of 8859 for spikes

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2792

Method: MS_TUNE Sample Name: Sample-0 Operator:
Comment:
Run Time: 03/02/99 20:45 Type: Unk Mode: INT Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	210400.
Stddev	2270.
%RSD	1.079

#1	207500.
#2	209300.
#3	210000.
#4	207200.
#5	212200.
#6	211700.
#7	209300.
#8	214600.
#9	211900.
#10	210100.

SP
3/2/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2793 TO 2823.

Analysis Report

Method: ACTINIDE Sample Name: cal blank

Operator: kjt

Comment: background

Run Time: 03/02/99 20:46 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2174	.2100	.2919	.3020	.1992
Stddev	.0239	.0208	.0260	.0421	.0408
%RSD	10.99	9.893	8.900	13.93	20.46

#1	.2292	.1834	.3331	.3056	.2690
#2	.2074	.2194	.3006	.3457	.1984
#3	.2120	.2210	.2807	.3375	.1852
#4	.1874	.1933	.2677	.2766	.1785
#5	.2508	.2331	.2773	.2449	.1652

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2233	.3295	.1970	.2086	.2208
Stddev	.0215	.0316	.0359	.0240	.0282
%RSD	9.612	9.605	18.20	11.51	12.75

#1	.2078	.3270	.2017	.2323	.2170
#2	.2134	.3547	.1593	.2104	.1803
#3	.2389	.3673	.2538	.2031	.2150
#4	.2528	.3064	.1755	.2261	.2350
#5	.2036	.2921	.1947	.1711	.2567

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2162	.2193
Stddev	.0356	.0160
%RSD	16.48	7.301

#1	.2231	.2384
#2	.2014	.2224
#3	.2568	.2299
#4	.1636	.2023
#5	.2360	.2036

Analysis Report

HNF-1663 REV. 0

03/02/99 20:50:30

page 2

Sample Name: cal blank Run Time: 03/02/99 20:46

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.33
Stddev	1.48
%RSD	1.3249

#1	109.06
#2	110.90
#3	111.62
#4	112.06
#5	112.99

Analysis Report

HNF-1663 REV. 0

03/02/99 20:55:06

page 1

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.010 ppm cal

Run Time: 03/02/99 20:51 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1587	.2106	796.7	931.6	1.353
Stddev	.0245	.0205	21.3	17.9	.048
%RSD	15.43	9.738	2.674	1.920	3.560

#1	.1931	.2289	762.2	927.0	1.415
#2	.1446	.1820	805.8	910.6	1.290
#3	.1751	.2318	790.5	921.0	1.326
#4	.1461	.2087	811.7	954.8	1.354
#5	.1345	.2018	813.3	944.5	1.381

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.660	.4959	.1653	.1535	.1388
Stddev	.046	.0303	.0311	.0229	.0347
%RSD	2.748	6.108	18.80	14.90	24.97

#1	1.724	.4717	.1848	.1765	.1848
#2	1.665	.5033	.2088	.1687	.1365
#3	1.641	.5409	.1314	.1185	.1262
#4	1.599	.4643	.1487	.1591	.0913
#5	1.671	.4992	.1526	.1449	.1552

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1647	.1367
Stddev	.0294	.0173
%RSD	17.85	12.63

#1	.2096	.1324
#2	.1553	.1472
#3	.1365	.1597
#4	.1774	.1148
#5	.1449	.1293

Analysis Report

HNF-1663 REV. 0

03/02/99 20:55:06

page 2

Sample Name: cal std Run Time: 03/02/99 20:51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	126.29
Stddev	3.59
%RSD	2.8402

#1	120.85
#2	124.52
#3	129.42
#4	127.78
#5	128.86

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.100 ppm

Run Time: 03/02/99 20:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2713	1.057	7912.	8793.	12.05
Stddev	.0243	.049	146.	186.	.20
%RSD	8.941	4.673	1.844	2.110	1.631

#1	.2968	1.132	7707.	8763.	11.93
#2	.2803	1.040	7878.	8754.	11.80
#3	.2836	1.019	7884.	8864.	12.05
#4	.2342	1.080	8100.	8537.	12.23
#5	.2615	1.014	7989.	9047.	12.26

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	15.23	1.983	.1560	.1377	.1535
Stddev	.36	.050	.0226	.0153	.0298
%RSD	2.368	2.499	14.50	11.14	19.40

#1	14.79	1.934	.1770	.1536	.1380
#2	15.14	2.033	.1834	.1284	.1231
#3	15.27	1.928	.1471	.1549	.1366
#4	15.18	2.025	.1353	.1223	.1770
#5	15.79	1.994	.1374	.1294	.1928

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1566	.1498
Stddev	.0079	.0192
%RSD	5.023	12.80

#1	.1588	.1588
#2	.1493	.1258
#3	.1549	.1444
#4	.1509	.1431
#5	.1691	.1770

Analysis Report

HNF-1663 REV. 0

Sample Name: cal std Run Time: 03/02/99 20:55

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	127.29
Stddev	.80
%RSD	.62839

#1	128.04
#2	127.23
#3	126.92
#4	128.08
#5	126.19

Analysis Report

HNF-1663 REV. 0

03/02/99 21:07:16

page 1

Method: ACTINIDE Sample Name: icv

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 21:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1771	.2734	850.9	14.85	1.432
Stddev	.0216	.0242	25.8	.38	.085
%RSD	12.18	8.847	3.035	2.557	5.944
#1	.1678	.2343	810.5	14.43	1.353
#2	.2116	.2665	880.5	14.89	1.525
#3	.1548	.2824	848.8	14.65	1.458
#4	.1692	.2900	851.3	14.82	1.490
#5	.1822	.2936	863.3	15.44	1.332

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1059.	961.3	26.51	.2862	.1890
Stddev	34.	25.1	.50	.0262	.0157
%RSD	3.205	2.616	1.900	9.156	8.332
#1	1002.	933.1	26.08	.2876	.1651
#2	1093.	949.3	26.52	.2748	.2033
#3	1064.	948.7	26.59	.3205	.1928
#4	1063.	984.7	26.06	.2981	.2014
#5	1071.	991.0	27.30	.2501	.1822

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1072.	.1883
Stddev	11.	.0305
%RSD	1.004	16.19
#1	1069.	.2077
#2	1081.	.2253
#3	1084.	.1521
#4	1072.	.1933
#5	1056.	.1631

Analysis Report

HNF-1663 REV. 0

03/02/99 21:07:16

page 2

Sample Name: icv Run Time: 03/02/99 21:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	123.19
Stddev	1.49
%RSD	1.2110

#1	125.17
#2	121.31
#3	122.74
#4	124.13
#5	122.60

Analysis Report

HNF-1663 REV. 0

03/02/99 21:18:54

page 1

Method: ACTINIDE Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/02/99 21:15 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1734	.1602	.3789	.2552	.1761
Stddev	.0161	.0354	.0172	.0105	.0284
%RSD	9.263	22.13	4.529	4.095	16.11

#1	.1676	.1886	.3588	.2541	.1807
#2	.2015	.1137	.3695	.2687	.1395
#3	.1701	.1944	.3888	.2619	.1701
#4	.1613	.1327	.3747	.2420	.2186
#5	.1663	.1715	.4028	.2495	.1715

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1707	.2651	.1818	.1719	.1766
Stddev	.0262	.0104	.0366	.0086	.0235
%RSD	15.36	3.928	20.11	4.990	13.34

#1	.1755	.2593	.1755	.1781	.1572
#2	.2041	.2687	.1809	.1576	.1524
#3	.1620	.2538	.2430	.1701	.1728
#4	.1795	.2810	.1613	.1769	.2081
#5	.1325	.2624	.1481	.1767	.1923

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1929	.1756
Stddev	.0216	.0120
%RSD	11.22	6.832

#1	.1833	.1729
#2	.1602	.1731
#3	.2025	.1836
#4	.2029	.1899
#5	.2157	.1585

Analysis Report

03/02/99 21:18:55

page 2

HNF-1663 REV. 0

Sample Name: icb Run Time: 03/02/99 21:15

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	127.22
Stddev	2.20
%RSD	1.7321

#1	127.26
#2	129.01
#3	123.44
#4	128.11
#5	128.28

Method: ACTINIDE Sample Name: blnk-prep

Operator: kjt

Comment: 11 df

Run Time: 03/02/99 21:19 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1814	.1860	.4922	.7370	.1962
Stddev	.0191	.0106	.0590	.0631	.0370
%RSD	10.52	5.701	11.99	8.564	18.87
#1	.1930	.1992	.5602	.8030	.2459
#2	.1793	.1851	.5176	.7894	.2255
#3	.1535	.1697	.4713	.7326	.1643
#4	.1771	.1883	.5088	.7140	.1771
#5	.2043	.1877	.4030	.6460	.1684

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2219	.4013	.1955	.1967	.1939
Stddev	.0157	.0566	.0142	.0256	.0201
%RSD	7.094	14.10	7.254	13.04	10.38
#1	.2334	.4824	.2054	.2303	.1899
#2	.2371	.4222	.2024	.1764	.2198
#3	.2047	.3717	.2074	.1670	.1643
#4	.2052	.3992	.1883	.2108	.1940
#5	.2291	.3313	.1739	.1988	.2015

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1894	.2144
Stddev	.0291	.0326
%RSD	15.36	15.21
#1	.1992	.2334
#2	.2053	.2487
#3	.1966	.1993
#4	.2080	.2249
#5	.1380	.1656

Sample Name: blk-prep Run Time: 03/02/99 21:19

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	117.09
Stddev	6.39
%RSD	5.4560

#1	107.10
#2	115.28
#3	123.77
#4	118.58
#5	120.75

Analysis Report

03/02/99 21:26:49

page 1

HNF-1663 REV.0

Method: ACTINIDE Sample Name: s98t002928

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 21:23 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1556	.1627	1.318	77.45	.1552
Stddev	.0081	.0227	.049	1.04	.0163
%RSD	5.235	13.94	3.702	1.342	10.50

#1	.1472	.1397	1.382	77.88	.1273
#2	.1644	.1993	1.325	78.77	.1644
#3	.1499	.1601	1.273	77.35	.1651
#4	.1642	.1493	1.267	75.92	.1543
#5	.1524	.1651	1.344	77.31	.1651

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2831	.2376	.1528	.1687	.1471
Stddev	.0236	.0235	.0256	.0168	.0206
%RSD	8.349	9.873	16.73	9.928	13.97

#1	.2894	.2670	.1572	.1946	.1747
#2	.2641	.2391	.1445	.1619	.1495
#3	.2591	.2490	.1829	.1651	.1372
#4	.3185	.2041	.1145	.1493	.1195
#5	.2845	.2286	.1651	.1727	.1549

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1606	.1728
Stddev	.0188	.0162
%RSD	11.69	9.399

#1	.1722	.1522
#2	.1644	.1594
#3	.1778	.1855
#4	.1593	.1891
#5	.1295	.1778

Analysis Report

03/02/99 21:26:50

page 2

HNF-1663 REV.0

Sample Name: S98t002928 Run Time: 03/02/99 21:23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	132.76
Stddev	1.41
%RSD	1.0634

#1	133.59
#2	133.81
#3	131.21
#4	133.94
#5	131.23

Method: ACTINIDE Sample Name: s98t002928_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 21:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1506	.1543	1.249	81.21	.1532
Stddev	.0344	.0227	.042	1.06	.0260
%RSD	22.83	14.73	3.361	1.310	16.99

#1	.1596	.1267	1.201	80.77	.1267
#2	.1076	.1333	1.212	82.20	.1486
#3	.1994	.1764	1.273	79.72	.1840
#4	.1558	.1711	1.256	81.09	.1762
#5	.1307	.1640	1.302	82.26	.1307

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3106	.2548	.1753	.1618	.1568
Stddev	.0431	.0195	.0354	.0223	.0166
%RSD	13.87	7.643	20.17	13.79	10.58

#1	.2584	.2508	.1469	.1900	.1672
#2	.2742	.2742	.1948	.1307	.1717
#3	.3578	.2275	.1891	.1610	.1661
#4	.3192	.2732	.1302	.1532	.1455
#5	.3433	.2485	.2152	.1742	.1332

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1440	.1497
Stddev	.0156	.0190
%RSD	10.87	12.68

#1	.1343	.1393
#2	.1435	.1666
#3	.1380	.1329
#4	.1711	.1736
#5	.1332	.1358

Sample Name: s98t002928_d Run Time: 03/02/99 21:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	130.54
Stddev	.61
%RSD	.46624

#1	131.57
#2	130.07
#3	130.41
#4	130.55
#5	130.10

Method: ACTINIDE Sample Name: s98t002929

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 21:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1433	.1594	1.562	47.18	.1659
Stddev	.0199	.0130	.020	.86	.0330
%RSD	13.90	8.187	1.307	1.825	19.87

#1	.1299	.1758	1.546	46.25	.1452
#2	.1387	.1544	1.565	46.76	.1780
#3	.1514	.1514	1.538	48.56	.1780
#4	.1735	.1450	1.590	47.28	.1217
#5	.1231	.1702	1.569	47.05	.2069

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2303	.2350	.1498	.1449	.1793
Stddev	.0346	.0213	.0264	.0135	.0183
%RSD	15.04	9.066	17.59	9.287	10.19

#1	.2267	.2369	.1248	.1325	.1579
#2	.2905	.2617	.1701	.1649	.1989
#3	.2125	.2151	.1806	.1408	.1620
#4	.2175	.2123	.1217	.1347	.1916
#5	.2042	.2488	.1519	.1519	.1859

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1438	.1755
Stddev	.0274	.0362
%RSD	19.08	20.61

#1	.1529	.1860
#2	.1152	.2146
#3	.1806	.1434
#4	.1528	.1321
#5	.1178	.2016

Sample Name: s98t002929 Run Time: 03/02/99 21:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	127.94
Stddev	1.98
%RSD	1.5502

#1	130.84
#2	127.36
#3	125.50
#4	128.72
#5	127.30

Analysis Report

HNF-1663 REV. 0

03/02/99 21:38:44

page 1

Method: ACTINIDE

Sample Name: s98t002929_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 21:35 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1551	.1628	1.646	49.23	.1602
Stddev	.0147	.0230	.066	1.04	.0130
%RSD	9.499	14.11	3.988	2.120	8.086

#1	.1709	.1326	1.538	48.65	.1505
#2	.1422	.1651	1.689	47.73	.1676
#3	.1691	.1947	1.707	49.45	.1794
#4	.1391	.1700	1.646	50.05	.1545
#5	.1543	.1517	1.649	50.26	.1492

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2550	.2452	.1899	.1695	.1567
Stddev	.0249	.0247	.0222	.0238	.0272
%RSD	9.764	10.07	11.70	14.04	17.36

#1	.2525	.2321	.1811	.1326	.1709
#2	.2692	.2642	.2235	.1600	.1600
#3	.2127	.2665	.1691	.1819	.1127
#4	.2730	.2549	.1751	.1803	.1545
#5	.2675	.2083	.2006	.1929	.1852

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1551	.1715
Stddev	.0238	.0096
%RSD	15.32	5.614

#1	.1505	.1556
#2	.1880	.1803
#3	.1332	.1717
#4	.1700	.1777
#5	.1337	.1723

HNF-1663 REV. 0

Sample Name: s98t002929_d Run Time: 03/02/99 21:35

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	130.21
Stddev	.75
%RSD	.57677

#1	130.70
#2	131.23
#3	130.09
#4	129.44
#5	129.61

HNF-1663 REV.0

Method: ACTINIDE Sample Name: S98t002930

Operator: kjt

Comment: 8241 df

Run Time: 03/02/99 21:39 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1534	.1662	.3247	997.4	.1599
Stddev	.0145	.0209	.0347	12.5	.0254
%RSD	9.486	12.59	10.69	1.255	15.89

#1	.1420	.1929	.3000	983.0	.1714
#2	.1537	.1643	.2783	1011.	.1511
#3	.1610	.1689	.3590	997.5	.1874
#4	.1370	.1344	.3532	1008.	.1687
#5	.1732	.1706	.3332	987.0	.1207

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.759	.4433	.1640	.1719	.1820
Stddev	.043	.0179	.0224	.0065	.0157
%RSD	2.413	4.036	13.66	3.807	8.609

#1	1.811	.4367	.1420	.1688	.1822
#2	1.725	.4240	.1988	.1670	.2041
#3	1.726	.4566	.1689	.1663	.1768
#4	1.800	.4322	.1634	.1766	.1608
#5	1.734	.4670	.1469	.1810	.1863

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1724	.1572
Stddev	.0223	.0175
%RSD	12.95	11.15

#1	.1554	.1661
#2	.1882	.1829
#3	.1425	.1452
#4	.1819	.1529
#5	.1942	.1391

Sample Name: s98t002930 Run Time: 03/02/99 21:39

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	126.00
Stddev	.99
%RSD .	.78348

#1	124.43
#2	125.78
#3	126.29
#4	126.48
#5	127.04

Method: ACTINIDE Sample Name: s98t002930_d Operator: kjt
 Comment: 8241 df
 Run Time: 03/02/99 21:43 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1662	.1764	.3232	626.0	.1662
Stddev	.0146	.0329	.0385	12.6	.0111
%RSD	8.796	18.67	11.90	2.017	6.672

#1	.1606	.1843	.3107	627.3	.1580
#2	.1735	.1864	.3391	616.0	.1838
#3	.1437	.1309	.3054	613.0	.1668
#4	.1815	.1607	.2800	628.7	.1555
#5	.1719	.2195	.3808	645.0	.1666

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.130	.3972	.1808	.1761	.1829
Stddev	.036	.0437	.0271	.0199	.0226
%RSD	3.157	10.99	14.98	11.32	12.38

#1	1.127	.3555	.2264	.2001	.1606
#2	1.129	.3909	.1735	.1527	.1812
#3	1.073	.3772	.1822	.1745	.1617
#4	1.161	.3914	.1633	.1918	.1996
#5	1.158	.4708	.1587	.1613	.2116

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1761	.1911
Stddev	.0163	.0251
%RSD	9.274	13.14

#1	.1764	.1501
#2	.1993	.1838
#3	.1565	.2079
#4	.1659	.2048
#5	.1825	.2089

Sample Name: s98t002930_d Run Time: 03/02/99 21:43

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	127.97
Stddev	1.61
%RSD	1.2554

#1	126.59
#2	128.76
#3	129.89
#4	128.58
#5	126.04

Method: ACTINIDE Sample Name: s98t002930_a

Operator: kjt

Comment: 8241 df

Run Time: 03/02/99 21:47 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1927	.3400	1571.	1025.	2.376
Stddev	.0167	.0333	35.	15.	.138
%RSD	8.674	9.783	2.216	1.462	5.796
#1	.1774	.3183	1515.	1000.	2.228
#2	.1805	.3745	1605.	1036.	2.387
#3	.1854	.3385	1591.	1038.	2.579
#4	.2038	.2977	1564.	1027.	2.271
#5	.2163	.3712	1581.	1026.	2.417

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2016.	1841.	50.21	.3851	.1883
Stddev	43.	38.	1.21	.0379	.0309
%RSD	2.130	2.042	2.409	9.846	16.39
#1	1940.	1777.	48.18	.3235	.1957
#2	2032.	1849.	50.02	.4095	.1751
#3	2045.	1847.	50.79	.3842	.2284
#4	2030.	1877.	50.92	.3862	.1448
#5	2033.	1854.	51.13	.4219	.1976

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2063.	.2172
Stddev	36.	.0203
%RSD	1.740	9.338
#1	2009.	.2139
#2	2086.	.2398
#3	2068.	.2257
#4	2050.	.1850
#5	2102.	.2216

Sample Name: s98t002930_a Run Time: 03/02/99 21:47

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	124.94
Stddev	1.63
%RSD	1.3074

#1	127.77
#2	123.73
#3	124.07
#4	124.30
#5	124.83

Analysis Report

03/02/99 21:58:38

page 1

actually the cv

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: icv

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 21:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1694	.2474	839.0	14.55	1.343
Stddev	.0128	.0224	5.9	.19	.052
%RSD	7.531	9.062	.7042	1.274	3.882
#1	.1664	.2781	835.9	14.74	1.310
#2	.1521	.2398	833.9	14.76	1.362
#3	.1696	.2338	846.5	14.42	1.390
#4	.1879	.2625	844.3	14.46	1.269
#5	.1711	.2230	834.6	14.37	1.384

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1077.	989.9	26.66	.2767	.1715
Stddev	9.	7.2	.19	.0350	.0176
%RSD	.8771	.7306	.7022	12.65	10.24
#1	1090.	989.0	26.43	.2339	.1793
#2	1075.	994.4	26.68	.3069	.1547
#3	1069.	980.0	26.93	.3032	.1978
#4	1083.	987.2	26.69	.2960	.1673
#5	1068.	999.0	26.55	.2437	.1581

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1101.	.1937
Stddev	7.	.0164
%RSD	.6569	8.482
#1	1101.	.1897
#2	1091.	.2089
#3	1099.	.2107
#4	1109.	.1879
#5	1107.	.1711

Analysis Report

03/02/99 21:58:39

page 2

Sample Name: icv

Run Time: 03/02/99 21:55

HNF-1663 REV. 0

Int. Std. Ir[193]
Line 193/pulse
Units Cts/S
Avg 129.07
Stdddev .63
%RSD .49140

#1 128.24
#2 129.27
#3 129.73
#4 129.52
#5 128.58

Method: ACTINIDE Sample Name: ccb

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 22:09 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1697	.1480	.2969	.2364	.1429
Stddev	.0169	.0175	.0183	.0347	.0303
%RSD	9.968	11.80	6.177	14.68	21.19

#1	.1937	.1274	.3210	.2043	.1247
#2	.1742	.1372	.2903	.2877	.1795
#3	.1468	.1442	.3014	.2464	.1311
#4	.1648	.1700	.3008	.2040	.1700
#5	.1692	.1614	.2708	.2395	.1093

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.1602	.3389	.1681	.1389	.1670
Stddev	.0133	.0204	.0102	.0386	.0174
%RSD	8.308	6.021	6.046	27.78	10.40

#1	.1672	.3555	.1565	.1937	.1406
#2	.1557	.3642	.1742	.1346	.1689
#3	.1389	.3329	.1625	.1573	.1782
#4	.1674	.3270	.1648	.0942	.1857
#5	.1718	.3150	.1822	.1146	.1614

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.1576	.1660
Stddev	.0159	.0165
%RSD	10.09	9.946

#1	.1459	.1486
#2	.1768	.1715
#3	.1389	.1835
#4	.1700	.1779
#5	.1562	.1484

Analysis Report

Sample Name: ccb Run Time: 03/02/99 22:09

HNF-1663 REV. 0

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	126.92
Stddev	.95
%RSD	.74792

#1	125.63
#2	126.31
#3	127.17
#4	127.44
#5	128.03

LABCORE Completed Worklist Report for Worklist# 27716

Analyst: kjt

Instrument: ICPMS1

Book#: 48857

Method: LA-506-101 Rev/Mod

A-D

49857

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	QMSPR-QC TC99	QC	2.05e-2	1.89e-02	92.195 % Recovery	
1 ICV	0	QMSPR-QC SN117	QC	1.925e-2	2.04e-02	105.974 % Recovery	
1 ICV	0	QMSPR-QC CS133	QC	2.00e-2	1.98e-02	99.000 % Recovery	
2 ICB	0	QMSPR-QC TC99	QC	1	3.00e-06	3.000e-006 ug/mL	
2 ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2 ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2 ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6	ug/mL	
2 ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2 ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3 BLNK-PREP	0	QMSPR-A1 TC99-A1	SOLID	1	5.39e-05	5.390e-005 ug/g	
3 BLNK-PREP	0	QMSPR-A1 SN117-A1	SOLID	1	7.17e-03	0.007 ug/g	
3 BLNK-PREP	0	QMSPR-A1 SN126-A1	SOLID	1	<1.65e-3	ug/g	
3 BLNK-PREP	0	QMSPR-A1 CS133-A1	SOLID	1	2.75e-05	2.750e-005 ug/g	
3 BLNK-PREP	0	QMSPR-A1 CS135-A1	SOLID	1	6.93e-05	6.930e-005 ug/g	
3 BLNK-PREP	0	QMSPR-A1 CS137-A1	SOLID	1	1.11e-04	1.110e-004 ug/g	
4 SAMPLE	S98T002713	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	1.412e+00	7.98e-002 ug/g
4 SAMPLE	S98T002713	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	5.982e+00	5.982 ug/g
4 SAMPLE	S98T002713	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	< 5.982e+00	5.982 ug/g
4 SAMPLE	S98T002713	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	3.210e+00	7.98e-002 ug/g
4 SAMPLE	S98T002713	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	4.598e+00	7.98e-002 ug/g
4 SAMPLE	S98T002713	0 A	QMSPR-A1 CS137-A1	SOLID	N/A	9.528e+00	7.98e-002 ug/g
5 DUP	S98T002713	0 A	QMSPR-A1 TC99-A1	SOLID	1.41e+00	1.01e+00	33.058 RPD
5 DUP	S98T002713	0 A	QMSPR-A1 SN117-A1	SOLID	<5.98e0	<5.87e0	RPD
5 DUP	S98T002713	0 A	QMSPR-A1 SN126-A1	SOLID	<5.98e0	<5.87e0	RPD
5 DUP	S98T002713	0 A	QMSPR-A1 CS133-A1	SOLID	3.21e+00	1.96e+00	48.356 RPD
5 DUP	S98T002713	0 A	QMSPR-A1 CS135-A1	SOLID	4.60e+00	2.82e+00	47.978 RPD
5 DUP	S98T002713	0 A	QMSPR-A1 CS137-A1	SOLID	9.53e+00	5.76e+00	49.313 RPD
6 SAMPLE	S98T002846	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	1.623e+00	8.20e-002 ug/g
6 SAMPLE	S98T002846	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	< 6.148e+00	6.149 ug/g
6 SAMPLE	S98T002846	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	< 6.148e+00	6.149 ug/g
6 SAMPLE	S98T002846	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	2.209e+00	8.20e-002 ug/g
6 SAMPLE	S98T002846	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	2.467e+00	8.20e-002 ug/g
6 SAMPLE	S98T002846	0 A	QMSPR-A1 CS137-A1	SOLID	N/A	4.513e+00	8.20e-002 ug/g
7 DUP	S98T002846	0 A	QMSPR-A1 TC99-A1	SOLID	1.62e+00	3.05e+00	61.242 RPD
7 DUP	S98T002846	0 A	QMSPR-A1 SN117-A1	SOLID	<6.15e0	<6.09e0	RPD
7 DUP	S98T002846	0 A	QMSPR-A1 SN126-A1	SOLID	<6.15e0	<6.09e0	RPD
7 DUP	S98T002846	0 A	QMSPR-A1 CS133-A1	SOLID	2.21e+00	3.82e+00	53.400 RPD
7 DUP	S98T002846	0 A	QMSPR-A1 CS135-A1	SOLID	2.47e+00	4.39e+00	55.977 RPD
7 DUP	S98T002846	0 A	QMSPR-A1 CS137-A1	SOLID	4.51e+00	7.98e+00	55.564 RPD
8 SAMPLE	S98T002740	0 A	QMSPR-A1 TC99-A1	SOLID	N/A	7.099e-01	8.21e-002 ug/g
8 SAMPLE	S98T002740	0 A	QMSPR-A1 SN117-A1	SOLID	N/A	< 6.156e+00	6.155 ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 27716

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
8 SAMPLE	S98T002740	0 A	QMSPR-A1 SN126-A1	SOLID	N/A	< 6.156e+00	6.155	ug/g
8 SAMPLE	S98T002740	0 A	QMSPR-A1 CS133-A1	SOLID	N/A	8.084e-01	8.21e-002	ug/g
8 SAMPLE	S98T002740	0 A	QMSPR-A1 CS135-A1	SOLID	N/A	4.022e-01	8.21e-002	ug/g
8 SAMPLE	S98T002740	0 A	QMSPR-A1 CS137-A1	SOLID	N/A	1.100e+00	8.21e-002	ug/g
9 DUP	S98T002740	0 A	QMSPR-A1 TC99-A1	SOLID	7.10e-01	7.99e-01	11.796	RPD
9 DUP	S98T002740	0 A	QMSPR-A1 SN117-A1	SOLID	<6.16e0	<5.85e0		RPD
9 DUP	S98T002740	0 A	QMSPR-A1 SN126-A1	SOLID	<6.16e0	<5.85e0		RPD
9 DUP	S98T002740	0 A	QMSPR-A1 CS133-A1	SOLID	8.08e-01	9.16e-01	12.529	RPD
9 DUP	S98T002740	0 A	QMSPR-A1 CS135-A1	SOLID	4.02e-01	4.72e-01	16.018	RPD
9 DUP	S98T002740	0 A	QMSPR-A1 CS137-A1	SOLID	1.10e+00	1.21e+00	9.524	RPD
10 SPK-POST	S98T002740	0 A	QMSPR-A1 TC99-A1	SOLID	0.0219	1.99e-02	90.868	% Recovery
10 SPK-POST	S98T002740	0 A	QMSPR-A1 SN117-A1	SOLID	0.0193	1.51e-02	78.238	% Recovery
10 SPK-POST	S98T002740	0 A	QMSPR-A1 CS133-A1	SOLID	0.0201	2.01e-02	100.000	% Recovery
11 CCV	0		QMSPR-QC TC99	QC	2.05e-2	1.87e-02	91.220	% Recovery
11 CCV	0		QMSPR-QC SN117	QC	1.925e-2	2.02e-02	104.935	% Recovery
11 CCV	0		QMSPR-QC CS133	QC	2.00e-2	1.95e-02	97.500	% Recovery
12 CCB	0		QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
12 CCB	0		QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
12 CCB	0		QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
12 CCB	0		QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27716

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

S98T002713 + S98T002846 will be rerun. High RPDs were not seen on 150-u or MSACT - maybe a dilution error.
Reanalyzed on WL # 27746. *[Signature]* 3/24/99

LABCORE Data Entry Template for Worklist# 27716

Analyst: KST Instrument: ICPMS1 Book# 48857

Method: LA-506-101 Rev/Mod A-0 49857

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSPR-QC	QC		
2	ICB		@MSPR-QC	QC		
3	BLNK-PREP		@MSPR-A1	SOLID		
4	SAMPLE	S98T002713 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
5	DUP	S98T002713 0 A	@MSPR-A1	SOLID		
6	SAMPLE	S98T002846 0 A	@MSPR-A1	SOLID	98000437	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
7	DUP	S98T002846 0 A	@MSPR-A1	SOLID		
8	~~SPK-POST~~	~~S98T002846 0 A~~	~~@MSPR-A1~~	~~SOLID~~		
9	SAMPLE	S98T002740 0 A	@MSPR-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
10	DUP	S98T002740 0 A	@MSPR-A1	SOLID		
11	SPK-POST	S98T002740 0 A	@MSPR-A1	SOLID		
12	CCV		@MSPR-QC	QC		
13	CCB		@MSPR-QC	QC		

Final page for worklist # 27716

KS Pharrndike 03-04-99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27716

S Type	Sample#	R A Test	Matrix	Group#	Project	Date
Analyst Signature		Date		Analyst Signature		

file pr0304a.txt
Blmk-Prep 1-10 11df
S98T002713 .025-10 201df
-d
2846
2846-d
2740
2740-d
2740-a
2740-a (Sn)
↓ ↓
.1ml of 44857
.25ml of 10 ppm Sn 48

Scans

pr 2713
pr 2846
pr 2740

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/04/99 12:42 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	175100.
Stddev	2492.
%RSD	1.423

#1	173700.
#2	177700.
#3	174800.
#4	171600.
#5	177200.
#6	178200.
#7	176100.
#8	171100.
#9	176900.
#10	174000.

file pro304a.txt
P-Bulk 1-10 11dF

SP 3/4/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2828 TO 2847.

Method: MS_TUNE

Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/04/99 12:43 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	174700.
Stddev	2474.
%RSD	1.416

#1	174500.
#2	172300.
#3	174700.
#4	169100.
#5	176600.
#6	173600.
#7	176800.
#8	175700.
#9	176800.
#10	176500.

Method: PRV

Sample Name: Blank

Operator:

Comment:

Run Time: 03/04/99 13:23 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.264	53.80	1.163	8.653	1.760
Stddev	.158	4.25	.168	.683	.280
%RSD	12.47	7.902	14.46	7.888	15.88

#1	1.467	51.00	.9413	8.222	1.772
#2	1.042	58.00	1.408	9.121	2.196
#3	1.306	58.50	1.164	8.687	1.448
#4	1.308	52.50	1.197	7.765	1.781
#5	1.195	49.00	1.107	9.471	1.603

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.090	2.590	3.277	14.11
Stddev	.284	.225	.317	.93
%RSD	13.59	8.673	9.687	6.592

#1	2.492	2.796	2.990	12.93
#2	2.111	2.562	3.012	14.98
#3	1.959	2.243	3.776	14.54
#4	2.171	2.783	3.312	13.30
#5	1.719	2.565	3.293	14.80

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.711
Stddev	.356
%RSD	2.0085

#1	18.061
#2	17.761
#3	17.612
#4	17.965
#5	17.157

Method: PRV

Sample Name: PRV CAL

Operator:

Comment:

Run Time: 03/04/99

13:26 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Cs[133]
Line	99/pulse	133/pulse
Units	Cts/S	Cts/S
Avg	20950.	19010.
Stddev	396.	596.
%RSD	1.889	3.137

#1	21520.	19920.
#2	21200.	19080.
#3	20690.	19100.
#4	20580.	18390.
#5	20760.	18570.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.634
Stddev	.484
%RSD	2.7432

#1	16.864
#2	17.493
#3	17.790
#4	17.927
#5	18.096

Method: PRV

Sample Name: PRV1 CAL

Operator:

Comment:

Run Time: 03/04/99 13:28 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Ru[102]	Te[125]	Cs[135]	Cs[137]
Line	102/pulse	125/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	592100.	1482.	6554.	11450.
Stddev	6560.	47.	157.	378.
%RSD	1.108	3.141	2.401	3.305

#1	595000.	1507.	6788.	11800.
#2	598200.	1432.	6370.	10800.
#3	589300.	1524.	6598.	11590.
#4	596000.	1430.	6462.	11510.
#5	581900.	1515.	6555.	11550.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	71130.
Stddev	2247.
%RSD	3.159

#1	73490.
#2	68270.
#3	72990.
#4	69430.
#5	71460.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.326
Stddev	.430
%RSD	2.4815

#1	17.144
#2	18.051
#3	17.030
#4	17.379
#5	17.025

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1063.5282	1.2635	1.0000000	03/04/99 13:27:29
Ru[102]	5920.3853	53.8000	1.0000000	03/04/99 13:29:56
Te[125]	14.8053	1.1634	1.0000000	03/04/99 13:29:56
Sn[117]	351.1060	8.6534	1.0000000	03/04/99 13:25:45
Sn[126]	351.4641	1.7599	1.0000000	03/04/99 13:25:45
Cs[133]	950.4922	2.0904	1.0000000	03/04/99 13:27:29
Cs[135]	992.7099	2.5897	1.0000000	03/04/99 13:29:56
Cs[137]	1021.8987	3.2766	1.0000000	03/04/99 13:29:56
Ba[138]	711.1303	14.1097	1.0000000	03/04/99 13:29:56

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV-SN CAL

Operator:

Comment:

Run Time: 03/04/99 13:31 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Sn[117]	Sn[126]
Line	117/pulse	126/pulse
Units	Cts/S	Cts/S
Avg	18540.	18540.
Stddev	132.	132.
%RSD	.7115	.7115

#1	18560.	18560.
#2	18570.	18570.
#3	18690.	18690.
#4	18330.	18330.
#5	18580.	18580.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.733
Stddev	.147
%RSD	.99589

#1	14.866
#2	14.892
#3	14.633
#4	14.721
#5	14.552

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1063.5282	1.2635	1.0000000	03/04/99 13:27:29
Ru[102]	5920.3853	53.8000	1.0000000	03/04/99 13:29:56
Te[125]	14.8053	1.1634	1.0000000	03/04/99 13:29:56
Sn[117]	962.8549	8.6534	1.0000000	03/04/99 13:32:51
Sn[126]	963.2130	1.7599	1.0000000	03/04/99 13:32:51
Cs[133]	950.4922	2.0904	1.0000000	03/04/99 13:27:29
Cs[135]	992.7099	2.5897	1.0000000	03/04/99 13:29:56
Cs[137]	1021.8987	3.2766	1.0000000	03/04/99 13:29:56
Ba[138]	711.1303	14.1097	1.0000000	03/04/99 13:29:56

Method: PRV

Sample Name: icv

Operator: kjt

Comment: msprv lcs

Run Time: 03/04/99 13:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.87	.0215	.0011	.9667	-.0001
Stddev	.24	.0085	.0184	.0963	.0003
%RSD	1.259	39.39	1692.	9.962	220.1

#1	18.79	.0353	.0092	.8241	-.0003
#2	18.86	.0223	.0237	.9265	.0001
#3	18.71	.0208	.0012	.9795	.0002
#4	19.27	.0160	-.0019	1.068	-.0004
#5	18.69	.0133	-.0268	1.036	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.75	.0068	.0110	.1049
Stddev	.34	.0003	.0005	.0038
%RSD	1.747	3.914	4.406	3.584

#1	19.65	.0069	.0108	.1053
#2	19.74	.0071	.0114	.1070
#3	19.65	.0064	.0117	.1066
#4	20.32	.0070	.0105	.1071
#5	19.39	.0068	.0107	.0983

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.465
Stddev	.208
%RSD	1.1909

#1	17.693
#2	17.507
#3	17.354
#4	17.168
#5	17.601

Method: PRV Sample Name: icv

Operator: kjt

Comment: msprv lcs sn

Run Time: 03/04/99 13:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0864	-.0002	.0112	20.38	.0001
Stddev	.0645	.0006	.0125	.18	.0001
%RSD	74.62	236.1	111.7	.9033	119.8
#1	.1936	-.0006	-.0033	20.66	.0002
#2	.0967	.0007	.0091	20.31	.0002
#3	.0630	-.0006	.0204	20.41	-.0001
#4	.0449	-.0002	.0271	20.15	.0003
#5	.0339	-.0004	.0028	20.38	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0124	.0272	.0464	.4252
Stddev	.0012	.0010	.0008	.0061
%RSD	10.05	3.619	1.676	1.431
#1	.0135	.0277	.0476	.4209
#2	.0137	.0265	.0463	.4221
#3	.0123	.0261	.0457	.4256
#4	.0116	.0274	.0457	.4216
#5	.0108	.0285	.0466	.4355

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.671
Stddev	.100
%RSD	.68283

#1	14.811
#2	14.644
#3	14.674
#4	14.693
#5	14.532

Method: PRV

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/04/99 13:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0030	.0196	-.0094	.0417	-.0001
Stddev	.0005	.0016	.0190	.0054	.0002
%RSD	15.19	8.165	202.6	12.90	152.6

#1	.0034	.0210	-.0241	.0471	.0001
#2	.0033	.0214	-.0025	.0479	.0000
#3	.0031	.0197	-.0240	.0386	-.0004
#4	.0023	.0185	-.0171	.0361	-.0004
#5	.0027	.0176	.0207	.0388	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0017	.0003	.0003	.0031
Stddev	.0003	.0001	.0004	.0010
%RSD	17.52	49.37	141.7	31.34

#1	.0015	.0005	.0005	.0038
#2	.0020	.0004	.0007	.0026
#3	.0019	.0003	-.0002	.0042
#4	.0013	.0002	-.0001	.0031
#5	.0018	.0001	.0007	.0017

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.313
Stddev	.210
%RSD	1.2150

#1	17.355
#2	17.301
#3	17.319
#4	17.591
#5	17.000

Method: PRV

Sample Name: prep-blank

Operator: kjt

Comment: 11 df

Run Time: 03/04/99 13:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0136	.0037	-.0032	.6776	.0000
Stddev	.0034	.0023	.0166	.0175	.000
%RSD	24.78	61.40	524.8	2.584	1022.

#1	.0187	.0062	.0126	.6947	-.0001
#2	.0148	.0053	.0057	.6882	.0005
#3	.0130	.0004	-.0295	.6639	-.0002
#4	.0120	.0031	-.0089	.6869	-.0002
#5	.0096	.0034	.0042	.6542	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0032	.0067	.0108	.1032
Stddev	.0003	.0004	.0012	.0041
%RSD	10.78	6.417	11.49	3.996

#1	.0031	.0073	.0103	.1072
#2	.0030	.0064	.0089	.1005
#3	.0037	.0062	.0121	.1081
#4	.0028	.0069	.0115	.1011
#5	.0032	.0066	.0112	.0992

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.921
Stddev	.307
%RSD	2.0584

#1	14.808
#2	15.231
#3	14.444
#4	15.026
#5	15.094

HNF-1663 REV. 0

Method: PRV

Sample Name: prep-blank

Operator: kjt

Comment: 11 df

Run Time: 03/04/99 13:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0049	.0023	-.0014	.6521	-.0001
Stddev	.0005	.0011	.0098	.0136	.0003
%RSD	10.48	49.59	702.6	2.085	185.4
#1	.0054	.0021	-.0143	.6416	-.0006
#2	.0054	.0025	.0116	.6521	.0001
#3	.0042	.0009	-.0014	.6666	-.0001
#4	.0047	.0040	-.0062	.6645	-.0001
#5	.0047	.0020	.0035	.6358	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0025	.0063	.0101	.0931
Stddev	.0001	.0009	.0005	.0066
%RSD	5.987	14.01	5.377	7.064
#1	.0024	.0065	.0098	.0860
#2	.0026	.0061	.0094	.0902
#3	.0025	.0077	.0108	.1018
#4	.0023	.0052	.0104	.0980
#5	.0026	.0062	.0102	.0893

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.130
Stddev	.206
%RSD	1.3618

#1	15.236
#2	15.358
#3	14.886
#4	14.939
#5	15.233

Method: PRV

Sample Name: s98t002740

Operator: kjt

Comment: 201 df

Run Time: 03/04/99 14:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0173	.0362	-.0083	.0333	.0000
Stddev	.0011	.0027	.0099	.0020	.000
%RSD	6.142	7.327	119.1	5.873	836.7

#1	.0168	.0372	-.0164	.0364	.0002
#2	.0191	.0375	-.0205	.0333	.0000
#3	.0163	.0345	.0022	.0329	.0000
#4	.0175	.0325	-.0001	.0309	-.0001
#5	.0168	.0392	-.0068	.0329	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0197	.0098	.0268	.0535
Stddev	.0005	.0006	.0011	.0006
%RSD	2.611	6.096	4.144	1.080

#1	.0190	.0104	.0285	.0537
#2	.0203	.0094	.0272	.0544
#3	.0197	.0093	.0260	.0531
#4	.0193	.0093	.0261	.0532
#5	.0199	.0105	.0260	.0531

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.222
Stddev	.136
%RSD	.70733

#1	19.004
#2	19.204
#3	19.237
#4	19.361
#5	19.303

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002740_d

Operator: kjt

Comment: 201 df

Run Time: 03/04/99 14:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0205	.0419	-.0037	.0366	.0001
Stddev	.0007	.0021	.0075	.0008	.0002
%RSD	3.322	4.894	204.3	2.318	205.3
#1	.0210	.0427	-.0115	.0379	.0005
#2	.0209	.0401	.0020	.0360	.0001
#3	.0203	.0397	-.0068	.0364	.0001
#4	.0207	.0421	.0064	.0358	-.0001
#5	.0193	.0448	-.0085	.0370	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0235	.0121	.0311	.0636
Stddev	.0012	.0004	.0017	.0024
%RSD	5.136	3.577	5.585	3.712
#1	.0244	.0123	.0335	.0673
#2	.0215	.0126	.0309	.0622
#3	.0234	.0117	.0298	.0637
#4	.0241	.0123	.0292	.0640
#5	.0241	.0115	.0321	.0610

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.013
Stddev	.508
%RSD	2.6731
#1	18.119
#2	19.284
#3	19.299
#4	19.077
#5	19.285

Method: PRV

Sample Name: s98t002740_a

Operator: kjt

Comment: 201 df

Run Time: 03/04/99 14:28 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.96	.0342	-.0127	3.449	.0000
Stddev	.15	.0032	.0088	.024	.000
%RSD	.7293	9.406	69.76	.7015	2381.

#1	20.18	.0347	-.0121	3.476	.0000
#2	19.99	.0314	.0022	3.461	.0001
#3	19.78	.0396	-.0166	3.426	-.0001
#4	19.93	.0325	-.0161	3.422	.0000
#5	19.91	.0329	-.0207	3.461	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	20.10	.0110	.0281	.0672
Stddev	.22	.0009	.0014	.0020
%RSD	1.083	8.422	5.021	2.955

#1	20.20	.0114	.0292	.0664
#2	20.20	.0107	.0298	.0659
#3	19.76	.0119	.0275	.0664
#4	20.01	.0095	.0262	.0664
#5	20.32	.0113	.0279	.0707

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.857
Stddev	.148
%RSD	.78438

#1	18.795
#2	18.820
#3	19.065
#4	18.932
#5	18.675

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002740_a

Operator: kjt

Comment: 201 df sn

Run Time: 03/04/99 14:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0188	.0343	-.0052	15.06	.0001
Stddev	.0004	.0013	.0145	.24	.0003
%RSD	2.126	3.873	280.0	1.579	247.9

#1	.0193	.0346	.0106	15.40	.0001
#2	.0187	.0357	-.0092	14.74	.0001
#3	.0192	.0337	-.0265	15.04	-.0001
#4	.0187	.0323	-.0063	15.03	-.0001
#5	.0183	.0352	.0056	15.10	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0227	.0117	.0292	.0694
Stddev	.0016	.0007	.0011	.0029
%RSD	7.260	5.588	3.706	4.114

#1	.0234	.0122	.0295	.0719
#2	.0217	.0117	.0294	.0724
#3	.0208	.0113	.0288	.0686
#4	.0227	.0108	.0306	.0686
#5	.0251	.0124	.0276	.0654

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.147
Stddev	.255
%RSD	1.4027

#1	17.802
#2	18.499
#3	18.155
#4	18.229
#5	18.050

HNF-1663 REV.0

Method: PRV

Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/04/99 14:40 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.73	.0013	-.0166	.9988	-.0003
Stddev	.11	.0008	.0049	.0919	.0002
%RSD	.5791	63.68	29.28	9.202	62.69
#1	18.87	.0010	-.0173	.8908	-.0004
#2	18.66	.0017	-.0089	.9346	-.0003
#3	18.60	.0021	-.0181	.9973	-.0004
#4	18.71	.0000	-.0223	1.047	-.0002
#5	18.81	.0016	-.0164	1.124	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.52	.0062	.0114	.1044
Stddev	.31	.0004	.0011	.0012
%RSD	1.572	5.625	9.893	1.138
#1	19.81	.0059	.0117	.1050
#2	19.17	.0068	.0096	.1032
#3	19.27	.0062	.0122	.1031
#4	19.51	.0063	.0124	.1052
#5	19.85	.0060	.0112	.1055

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.189
Stddev	.234
%RSD	1.2840

#1	18.175
#2	18.417
#3	18.431
#4	17.998
#5	17.922

HNF-1663 REV. 0

Method: PRV

Sample Name: ccv

Operator: kjt

Comment: lcs sn

Run Time: 03/04/99 14:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0732	-.0009	.0099	20.21	.0000
Stddev	.0590	.0003	.0172	.09	.0004
%RSD	80.58	34.27	172.9	.4592	8201.

#1	.1715	-.0007	-.0015	20.18	.0001
#2	.0827	-.0008	-.0027	20.35	.0002
#3	.0502	-.0011	.0175	20.09	-.0005
#4	.0345	-.0013	-.0005	20.21	.0004
#5	.0270	-.0006	.0368	20.22	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0119	.0276	.0455	.4155
Stddev	.0006	.0008	.0012	.0039
%RSD	4.960	2.869	2.704	.9364

#1	.0123	.0276	.0472	.4125
#2	.0126	.0269	.0458	.4134
#3	.0120	.0273	.0448	.4175
#4	.0116	.0275	.0439	.4126
#5	.0111	.0290	.0459	.4214

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.128
Stddev	.144
%RSD	.94895

#1	15.331
#2	15.138
#3	15.111
#4	15.134
#5	14.926

Method: PRV Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/04/99 14:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	.0024	-.0042	.0230	-.0002
Stddev	.0002	.0015	.0090	.0017	.0002
%RSD	48.81	64.71	216.5	7.355	109.0
#1	.0003	.0037	-.0057	.0246	-.0003
#2	.0003	.0027	-.0095	.0250	.0001
#3	.0006	.0037	.0106	.0220	-.0003
#4	.0007	.0000	-.0130	.0224	-.0003
#5	.0002	.0017	-.0032	.0211	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0017	-.0001	.0004	.0001
Stddev	.0002	.0001	.0003	.0012
%RSD	11.68	158.6	75.29	835.9
#1	.0020	.0001	.0003	-.0013
#2	.0017	-.0001	.0003	-.0002
#3	.0015	-.0002	.0010	.0010
#4	.0017	.0000	.0002	.0017
#5	.0015	-.0001	.0004	-.0004

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.526
Stddev	.083
%RSD	.47446

#1	17.612
#2	17.610
#3	17.424
#4	17.503
#5	17.480

LABCORE Completed Worklist Report for Worklist# 27717

Analyst: kjt Instrument: ICPMS1

Book#: 48857
49857

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK-PREP	0		QMSPR-A1 TC99-A1 SOLID		1	1.76e-04	1.760e-004	ug/g
1 BLNK-PREP	0		QMSPR-A1 SN117-A1 SOLID		1	5.18e-03	0.005	ug/g
1 BLNK-PREP	0		QMSPR-A1 SN126-A1 SOLID		1	<1.65e-3		ug/g
1 BLNK-PREP	0		QMSPR-A1 CS133-A1 SOLID		1	<2.20e-5		ug/g
1 BLNK-PREP	0		QMSPR-A1 CS135-A1 SOLID		1	4.84e-05	4.840e-005	ug/g
1 BLNK-PREP	0		QMSPR-A1 CS137-A1 SOLID		1	8.80e-05	8.800e-005	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 TC99-A1 SOLID	N/A		2.008e-01	8.20e-002	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 SN117-A1 SOLID	N/A <		6.148e+00	6.149	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 SN126-A1 SOLID	N/A <		6.148e+00	6.149	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 CS133-A1 SOLID	N/A		2.746e-01	8.20e-002	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 CS135-A1 SOLID	N/A		8.607e-02	8.20e-002	ug/g
2 SAMPLE	S98T002928	0 A	QMSPR-A1 CS137-A1 SOLID	N/A		2.787e-01	8.20e-002	ug/g
3 DUP	S98T002928	0 A	QMSPR-A1 TC99-A1 SOLID		2.01e-01	2.14e-01	6.265	RPD
3 DUP	S98T002928	0 A	QMSPR-A1 SN117-A1 SOLID		<6.15e0	<5.73e0		RPD
3 DUP	S98T002928	0 A	QMSPR-A1 SN126-A1 SOLID		<6.15e0	<5.73e0		RPD
3 DUP	S98T002928	0 A	QMSPR-A1 CS133-A1 SOLID		2.75e-01	2.63e-01	4.461	RPD
3 DUP	S98T002928	0 A	QMSPR-A1 CS135-A1 SOLID		8.61e-02	9.16e-02	6.190	RPD
3 DUP	S98T002928	0 A	QMSPR-A1 CS137-A1 SOLID		2.79e-01	2.83e-01	1.423	RPD
4 SAMPLE	S98T002929	0 A	QMSPR-A1 TC99-A1 SOLID	N/A		9.330e-01	7.91e-002	ug/g
4 SAMPLE	S98T002929	0 A	QMSPR-A1 SN117-A1 SOLID	N/A <		5.930e+00	5.930	ug/g
4 SAMPLE	S98T002929	0 A	QMSPR-A1 SN126-A1 SOLID	N/A <		5.930e+00	5.930	ug/g
4 SAMPLE	S98T002929	0 A	QMSPR-A1 CS133-A1 SOLID	N/A		1.218e+00	7.91e-002	ug/g
4 SAMPLE	S98T002929	0 A	QMSPR-A1 CS135-A1 SOLID	N/A		4.705e-01	7.91e-002	ug/g
4 SAMPLE	S98T002929	0 A	QMSPR-A1 CS137-A1 SOLID	N/A		6.286e-01	7.91e-002	ug/g
5 DUP	S98T002929	0 A	QMSPR-A1 TC99-A1 SOLID		9.33e-01	9.44e-01	1.172	RPD
5 DUP	S98T002929	0 A	QMSPR-A1 SN117-A1 SOLID		<5.93e0	<6.13e0		RPD
5 DUP	S98T002929	0 A	QMSPR-A1 SN126-A1 SOLID		<5.93e0	<6.13e0		RPD
5 DUP	S98T002929	0 A	QMSPR-A1 CS133-A1 SOLID		1.22e+00	1.23e+00	0.816	RPD
5 DUP	S98T002929	0 A	QMSPR-A1 CS135-A1 SOLID		4.70e-01	4.66e-01	0.855	RPD
5 DUP	S98T002929	0 A	QMSPR-A1 CS137-A1 SOLID		6.29e-01	6.21e-01	1.280	RPD
6 SAMPLE	S98T002930	0 A	QMSPR-A1 TC99-A1 SOLID	N/A <		3.204e+00	3.204	ug/g
6 SAMPLE	S98T002930	0 A	QMSPR-A1 SN117-A1 SOLID	N/A <		2.403e+02	240.300	ug/g
6 SAMPLE	S98T002930	0 A	QMSPR-A1 SN126-A1 SOLID	N/A <		2.403e+02	240.300	ug/g
6 SAMPLE	S98T002930	0 A	QMSPR-A1 CS133-A1 SOLID	N/A <		3.204e+00	3.204	ug/g
6 SAMPLE	S98T002930	0 A	QMSPR-A1 CS135-A1 SOLID	N/A		7.690e+00	3.204	ug/g
6 SAMPLE	S98T002930	0 A	QMSPR-A1 CS137-A1 SOLID	N/A		1.634e+01	3.204	ug/g
7 DUP	S98T002930	0 A	QMSPR-A1 TC99-A1 SOLID		<3.20e0	<3.26e0		RPD
7 DUP	S98T002930	0 A	QMSPR-A1 SN117-A1 SOLID		<2.40e2	<2.44e2		RPD
7 DUP	S98T002930	0 A	QMSPR-A1 SN126-A1 SOLID		<2.40e2	<2.44e2		RPD
7 DUP	S98T002930	0 A	QMSPR-A1 CS133-A1 SOLID		<3.20e0	<3.26e0		RPD
7 DUP	S98T002930	0 A	QMSPR-A1 CS135-A1 SOLID		7.69e+00	6.35e+00	19.088	RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27717

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
7	DUP	S98T002930	0	A	@MSPR-A1 CS137-A1	SOLID	1.63e+01	1.30e+01	22.526	RPD
8	CCV		0		@MSPR-QC TC99	QC	0.0205	1.89e-02	92.195 %	Recovery
8	CCV		0		@MSPR-QC SN117	QC	0.01925	2.15e-02	111.688 %	Recovery
8	CCV		0		@MSPR-QC CS133	QC	0.02000	1.96e-02	98.000 %	Recovery
9	CCB		0		@MSPR-QC TC99	QC	1	<2.00e-6		ug/mL
9	CCB		0		@MSPR-QC SN117	QC	1	<1.50e-4		ug/mL
9	CCB		0		@MSPR-QC SN126	QC	1	<1.50e-4		ug/mL
9	CCB		0		@MSPR-QC CS133	QC	1	<2.00e-6		ug/mL
9	CCB		0		@MSPR-QC CS135	QC	1	<2.00e-6		ug/mL
9	CCB		0		@MSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27717

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sal H. Fang 03/08/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27717

Analyst: KSR Instrument: ICPMS1 Book# 48857

Method: LA-506-101 Rev/Mod A-0 49857

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK-PREP			@MSPR-A1 SOLID			
2 SAMPLE	S98T002928 0 A		@MSPR-A1 SOLID		98000490 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
3 DUP	S98T002928 0 A		@MSPR-A1 SOLID			
4 SAMPLE	S98T002929 0 A		@MSPR-A1 SOLID		98000490 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
5 DUP	S98T002929 0 A		@MSPR-A1 SOLID			
6 SAMPLE	S98T002930 0 A		@MSPR-A1 SOLID		98000490 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
7 DUP	S98T002930 0 A		@MSPR-A1 SOLID			
8 CCV			@MSPR-QC QC			
9 CCB			@MSPR-QC QC			

Final page for worklist # 27717

KSR 3-7-99
Signature Date

file c307a.txt
Signature Date
retest new dilutions
S98T002928 .05-10 201df
-2 - -

Blk-Prep 1-10 11df
S98T002928 .05-10 201df
2428-d
2429
2429-d
2430 .05-10-250-10 8241df
Data Entry Comments: 2430-d

Scans pr 27717
pr 2928
pr 2929
pr 2930

Spokes .1ml of msprv spike 44857
.25ml of 10ppm (99019-1) Sn

Method: PRV

Sample Name: blk-prep

Operator: kjt

Comment: 11 df

Run Time: 03/07/99 11:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0160	-.0002	.0069	.4705	.0002
Stddev	.0052	.0006	.0137	.0163	.0002
%RSD	32.77	262.9	199.2	3.461	98.66

#1	.0211	-.0009	-.0016	.4431	.0001
#2	.0212	.0001	.0251	.4699	.0002
#3	.0153	-.0004	.0098	.4835	.0004
#4	.0130	-.0005	.0117	.4817	-.0001
#5	.0092	.0006	-.0108	.4743	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0011	.0044	.0080	.0713
Stddev	.0004	.0007	.0004	.0014
%RSD	37.95	14.82	5.320	2.018

#1	.0015	.0048	.0074	.0728
#2	.0011	.0052	.0080	.0716
#3	.0009	.0035	.0086	.0699
#4	.0016	.0041	.0077	.0698
#5	.0006	.0044	.0081	.0726

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.747
Stddev	.346
%RSD	2.1943

#1	16.120
#2	15.510
#3	15.933
#4	15.274
#5	15.901

Uploaded & Validated by:
Saul M. Paving
03/08/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2851 TO 2862

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002928

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0049	.0048	-.0072	.0243	.0004
Stddev	.0001	.0011	.0098	.0008	.0001
%RSD	2.967	23.94	137.3	3.495	34.21

#1	.0049	.0044	-.0140	.0253	.0003
#2	.0049	.0059	-.0110	.0247	.0003
#3	.0046	.0031	.0101	.0230	.0003
#4	.0050	.0048	-.0121	.0240	.0006
#5	.0048	.0057	-.0087	.0245	.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0067	.0021	.0068	.0069
Stddev	.0008	.0002	.0005	.0011
%RSD	11.29	10.18	7.375	16.05

#1	.0061	.0021	.0061	.0062
#2	.0060	.0022	.0068	.0063
#3	.0065	.0024	.0071	.0058
#4	.0077	.0022	.0074	.0084
#5	.0072	.0018	.0065	.0077

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.034
Stddev	.084
%RSD	.46535

#1	18.010
#2	18.171
#3	17.956
#4	17.984
#5	18.047

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002928_d

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0056	.0046	-.0088	.0504	.0004
Stddev	.0005	.0011	.0127	.0014	.0003
%RSD	8.186	23.07	143.7	2.855	72.23

#1	.0061	.0048	.0080	.0498	.0003
#2	.0057	.0044	-.0252	.0482	.0005
#3	.0051	.0028	-.0162	.0507	.0002
#4	.0051	.0054	-.0027	.0519	.0008
#5	.0059	.0054	-.0079	.0514	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0069	.0024	.0074	.0086
Stddev	.0006	.0003	.0002	.0005
%RSD	8.113	13.51	3.158	5.734

#1	.0077	.0021	.0075	.0095
#2	.0072	.0024	.0073	.0085
#3	.0064	.0020	.0073	.0084
#4	.0065	.0026	.0076	.0083
#5	.0066	.0028	.0070	.0085

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.992
Stddev	.089
%RSD	.49315

#1	18.011
#2	18.050
#3	18.090
#4	17.940
#5	17.867

Method: PRV

Sample Name: s98t002929

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0236	.0166	-.0146	.0318	.0003
Stddev	.0009	.0013	.0163	.0015	.0002
%RSD	3.854	7.785	111.5	4.623	75.18
#1	.0230	.0171	-.0087	.0323	.0000
#2	.0243	.0176	-.0291	.0309	.0004
#3	.0224	.0180	-.0287	.0298	.0003
#4	.0239	.0151	.0102	.0333	.0005
#5	.0246	.0154	-.0168	.0328	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0308	.0119	.0159	.0061
Stddev	.0009	.0003	.0004	.0010
%RSD	3.027	2.598	2.587	15.84
#1	.0314	.0121	.0158	.0072
#2	.0308	.0116	.0161	.0047
#3	.0307	.0116	.0165	.0057
#4	.0294	.0122	.0154	.0060
#5	.0319	.0121	.0158	.0066

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.080
Stddev	.121
%RSD	.67086

#1	18.053
#2	18.152
#3	18.010
#4	17.937
#5	18.246

Method: PRV

Sample Name: s98t002929_d

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0231	.0168	-.0151	.0282	.0004
Stddev	.0007	.0002	.0151	.0007	.0003
%RSD	2.980	1.134	100.1	2.426	73.04
#1	.0224	.0171	-.0092	.0282	.0007
#2	.0237	.0166	-.0038	.0282	.0003
#3	.0239	.0167	-.0188	.0288	.0006
#4	.0228	.0169	-.0038	.0286	.0000
#5	.0225	.0167	-.0398	.0271	.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0301	.0114	.0152	.0047
Stddev	.0008	.0005	.0008	.0012
%RSD	2.782	3.998	5.337	24.82
#1	.0315	.0120	.0150	.0065
#2	.0303	.0117	.0152	.0038
#3	.0296	.0110	.0143	.0052
#4	.0299	.0112	.0165	.0037
#5	.0293	.0110	.0150	.0042

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.176
Stddev	.092
%RSD	.50430

#1	18.163
#2	18.186
#3	18.302
#4	18.187
#5	18.044

Method: PRV

Sample Name: s98t002930

Operator: kjt

Comment: 8241 df

Run Time: 03/07/99 11:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0069	-.0078	.0005	.0002
Stddev	.0003	.0019	.0126	.0007	.0002
%RSD	37.50	27.97	162.1	129.5	119.0
#1	.0014	.0041	-.0134	.0004	.0001
#2	.0005	.0083	-.0029	.0006	.0005
#3	.0009	.0078	.0022	.0016	.0000
#4	.0007	.0086	.0025	-.0002	.0003
#5	.0008	.0057	-.0272	.0002	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0011	.0048	.0102	.0791
Stddev	.0004	.0004	.0004	.0011
%RSD	39.07	8.122	3.471	1.349
#1	.0015	.0052	.0101	.0791
#2	.0012	.0041	.0101	.0804
#3	.0007	.0050	.0107	.0780
#4	.0006	.0048	.0102	.0782
#5	.0014	.0047	.0097	.0800

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.418
Stddev	.242
%RSD	1.3135

#1	18.779
#2	18.382
#3	18.445
#4	18.386
#5	18.101

HNF-1663 REV.0

Method: PRV

Sample Name: s98t002930_d

Operator: kjt

Comment: 8241 df

Run Time: 03/07/99 11:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0037	-.0045	.0003	.0002
Stddev	.0004	.0012	.0066	.0005	.0001
%RSD	49.59	32.48	145.7	144.6	66.70

#1	.0003	.0030	.0063	.0008	.0001
#2	.0008	.0023	-.0075	.0005	.0003
#3	.0013	.0032	-.0077	.0007	.0001
#4	.0006	.0044	-.0106	.0000	.0004
#5	.0008	.0053	-.0032	-.0003	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0011	.0039	.0080	.0624
Stddev	.0002	.0002	.0003	.0023
%RSD	14.48	4.385	4.016	3.698

#1	.0010	.0038	.0080	.0597
#2	.0009	.0040	.0084	.0622
#3	.0010	.0038	.0078	.0659
#4	.0013	.0036	.0076	.0614
#5	.0012	.0040	.0081	.0631

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.251
Stddev	.354
%RSD	1.9375

#1	18.696
#2	18.189
#3	17.820
#4	18.511
#5	18.040

Method: PRV

Sample Name: s98t0022928

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0045	.0044	-.0162	.0228	.0003
Stddev	.0004	.0011	.0143	.0014	.0003
%RSD	9.610	26.40	88.37	6.129	87.23
#1	.0044	.0044	-.0174	.0228	.0006
#2	.0042	.0060	-.0022	.0245	.0006
#3	.0047	.0038	-.0365	.0230	.0002
#4	.0051	.0029	-.0219	.0206	.0001
#5	.0040	.0047	-.0030	.0231	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0057	.0022	.0073	.0067
Stddev	.0003	.0002	.0006	.0008
%RSD	6.042	8.725	8.222	11.24
#1	.0051	.0019	.0062	.0072
#2	.0058	.0023	.0073	.0076
#3	.0060	.0022	.0074	.0064
#4	.0059	.0022	.0078	.0056
#5	.0055	.0024	.0076	.0066

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	18.105
Stddev	.127
%RSD	.70222

#1	17.944
#2	18.217
#3	18.199
#4	18.172
#5	17.992

Method: PRV

Sample Name: s98t0022928_d

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 11:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0054	.0036	-.0004	.0487	.0001
Stddev	.0005	.0009	.0049	.0018	.0002
%RSD	8.458	25.02	1298.	3.778	125.0

#1	.0055	.0028	.0069	.0492	.0000
#2	.0062	.0043	-.0013	.0509	.0000
#3	.0050	.0025	-.0068	.0472	.0001
#4	.0053	.0038	-.0002	.0465	.0004
#5	.0052	.0046	-.0005	.0499	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0060	.0028	.0078	.0078
Stddev	.0005	.0005	.0005	.0013
%RSD	8.610	16.51	5.821	16.76

#1	.0058	.0026	.0079	.0095
#2	.0063	.0028	.0082	.0071
#3	.0068	.0037	.0082	.0064
#4	.0057	.0027	.0072	.0089
#5	.0055	.0025	.0074	.0072

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.142
Stddev	.098
%RSD	.53952

#1	18.212
#2	18.040
#3	18.036
#4	18.177
#5	18.245

HNF-1663 REV. 0

Method: PRV

Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/07/99 11:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.86	.0014	.0003	1.428	.0000
Stddev	.21	.0009	.0135	.096	.0002
%RSD	1.093	64.79	3953.	6.749	1395.
#1	18.79	.0028	-.0131	1.289	-.0001
#2	19.17	.0014	.0208	1.398	-.0002
#3	18.90	.0010	-.0011	1.417	.0001
#4	18.60	.0012	.0051	1.496	.0003
#5	18.86	.0004	-.0100	1.539	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.62	.0075	.0120	.1138
Stddev	.11	.0003	.0004	.0024
%RSD	.5432	3.660	3.700	2.088
#1	19.57	.0079	.0114	.1113
#2	19.61	.0075	.0126	.1120
#3	19.66	.0071	.0119	.1129
#4	19.76	.0074	.0122	.1160
#5	19.48	.0074	.0118	.1166

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.000
Stddev	.151
%RSD	.84051

#1	18.244
#2	18.007
#3	17.991
#4	17.841
#5	17.917

HNF-1663 REV. 0

Method: PRV Sample Name: ccv

Operator: Kjt

Comment: 19.25 ppb Sn

Run Time: 03/07/99 11:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1786	.0009	.0032	21.53	.0001
Stddev	.1852	.0006	.0170	.31	.0003
%RSD	103.7	64.82	526.6	1.421	247.7
#1	.4857	.0010	.0045	21.98	-.0003
#2	.2176	.0005	.0009	21.32	.0004
#3	.0928	.0014	.0308	21.21	-.0001
#4	.0569	.0015	-.0055	21.68	.0002
#5	.0401	.0001	-.0145	21.46	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0120	.0289	.0470	.4358
Stddev	.0025	.0008	.0008	.0065
%RSD	20.86	2.739	1.702	1.492
#1	.0159	.0301	.0459	.4354
#2	.0130	.0284	.0472	.4302
#3	.0103	.0293	.0469	.4375
#4	.0110	.0286	.0470	.4301
#5	.0098	.0282	.0481	.4459

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	15.456
Stddev	.168
%RSD	1.0878

#1	15.402
#2	15.658
#3	15.606
#4	15.268
#5	15.348

Method: PRV Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/07/99 12:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	.0013	-.0112	.0037	.0002
Stddev	.0001	.0012	.0130	.0007	.0002
%RSD	32.99	96.36	115.6	17.68	121.1

#1	-.0002	.0011	.0115	.0035	.0001
#2	-.0002	.0007	-.0141	.0037	.0002
#3	-.0003	.0031	-.0199	.0039	.0003
#4	-.0002	-.0002	-.0144	.0029	-.0001
#5	-.0004	.0016	-.0192	.0047	.0005

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0002	-.0001	.0000	.0008
Stddev	.0004	.0002	.0003	.0009
%RSD	236.0	201.3	992.6	123.2

#1	.0004	-.0003	-.0004	-.0007
#2	.0004	-.0002	.0002	.0012
#3	.0004	-.0001	.0000	.0008
#4	-.0005	.0002	.0000	.0018
#5	.0002	.0000	.0004	.0007

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.517
Stddev	.122
%RSD	.69436

#1	17.355
#2	17.561
#3	17.609
#4	17.636
#5	17.425

LABCORE Completed Worklist Report for Worklist# 27718

Analyst: bjg

Instrument: ICPMS1

Book#: 48857, 49857

Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSPR-QC TC99	QC	0.0205	1.92e-02	93.659 % Recovery	
1	ICV	0	QMSPR-QC SN117	QC	0.01856	2.14e-02	115.302 % Recovery	
1	ICV	0	QMSPR-QC CS133	QC	0.0200	1.98e-02	99.000 % Recovery	
2	ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
2	ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
2	ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL
3	BLNK-PREP	0	QMSPR-A1 TC99-A1	SOLID	1	1.47e-04	1.470e-004	ug/g
3	BLNK-PREP	0	QMSPR-A1 SN117-A1	SOLID	1	4.41e-03	0.004	ug/g
3	BLNK-PREP	0	QMSPR-A1 SN126-A1	SOLID	1	<1.50e-3		ug/g
3	BLNK-PREP	0	QMSPR-A1 CS133-A1	SOLID	1	<2.00e-5		ug/g
3	BLNK-PREP	0	QMSPR-A1 CS135-A1	SOLID	1	4.30e-05	4.300e-005	ug/g
3	BLNK-PREP	0	QMSPR-A1 CS137-A1	SOLID	1	6.80e-05	6.800e-005	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 TC99-A1	SOLID	N/A	2.686e-01	7.90e-002	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 SN117-A1	SOLID	N/A <	5.926e+00	5.925	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 SN126-A1	SOLID	N/A <	5.926e+00	5.925	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 CS133-A1	SOLID	N/A	2.844e-01	7.90e-002	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 CS135-A1	SOLID	N/A	1.383e-01	7.90e-002	ug/g
4	SAMPLE	S98T003394	0 A QMSPR-A1 CS137-A1	SOLID	N/A	4.425e-01	7.90e-002	ug/g
5	DUP	S98T003394	0 A QMSPR-A1 TC99-A1	SOLID	2.69e-01	3.10e-01	14.162	RPD
5	DUP	S98T003394	0 A QMSPR-A1 SN117-A1	SOLID	<5.93e0	<6.20e0		RPD
5	DUP	S98T003394	0 A QMSPR-A1 SN126-A1	SOLID	<5.93e0	<6.20e0		RPD
5	DUP	S98T003394	0 A QMSPR-A1 CS133-A1	SOLID	2.84e-01	3.35e-01	16.478	RPD
5	DUP	S98T003394	0 A QMSPR-A1 CS135-A1	SOLID	1.38e-01	1.24e-01	10.687	RPD
5	DUP	S98T003394	0 A QMSPR-A1 CS137-A1	SOLID	4.43e-01	2.68e-01	49.226	RPD
6	SAMPLE	S98T003399	0 A QMSPR-A1 TC99-A1	SOLID	N/A <	8.143e+00	8.143	ug/g
6	SAMPLE	S98T003399	0 A QMSPR-A1 SN117-A1	SOLID	N/A <	6.107e+02	610.700	ug/g
6	SAMPLE	S98T003399	0 A QMSPR-A1 SN126-A1	SOLID	N/A <	6.107e+02	610.700	ug/g
6	SAMPLE	S98T003399	0 A QMSPR-A1 CS133-A1	SOLID	N/A <	8.143e+00	8.143	ug/g
6	SAMPLE	S98T003399	0 A QMSPR-A1 CS135-A1	SOLID	N/A	2.809e+01	8.143	ug/g
6	SAMPLE	S98T003399	0 A QMSPR-A1 CS137-A1	SOLID	N/A	5.660e+01	8.143	ug/g
7	DUP	S98T003399	0 A QMSPR-A1 TC99-A1	SOLID	<8.14e0	<7.79e0		RPD
7	DUP	S98T003399	0 A QMSPR-A1 SN117-A1	SOLID	<6.11e2	<5.84e2		RPD
7	DUP	S98T003399	0 A QMSPR-A1 SN126-A1	SOLID	<6.11e2	<5.84e2		RPD
7	DUP	S98T003399	0 A QMSPR-A1 CS133-A1	SOLID	<8.14e0	<7.79e0		RPD
7	DUP	S98T003399	0 A QMSPR-A1 CS135-A1	SOLID	2.81e+01	3.08e+01	9.168	RPD
7	DUP	S98T003399	0 A QMSPR-A1 CS137-A1	SOLID	5.66e+01	6.27e+01	10.226	RPD
8	SAMPLE	S98T003389	0 A QMSPR-A1 TC99-A1	SOLID	N/A	4.656e-01	8.24e-002	ug/g
8	SAMPLE	S98T003389	0 A QMSPR-A1 SN117-A1	SOLID	N/A <	6.181e+00	6.181	ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27718

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
17 DUP	S98T002608	0 A	QMSPR-A1 CS133-A1 SOLID		<8.03e0	<7.85e0		RPD
17 DUP	S98T002608	0 A	QMSPR-A1 CS135-A1 SOLID		2.93e+01	2.67e+01	9.286	RPD
17 DUP	S98T002608	0 A	QMSPR-A1 CS137-A1 SOLID		6.30e+01	5.50e+01	13.559	RPD
18 SAMPLE	S98T002609	0 A	QMSPR-A1 TC99-A1 SOLID		N/A	<	8.259e+00	ug/g
18 SAMPLE	S98T002609	0 A	QMSPR-A1 SN117-A1 SOLID		N/A	<	6.194e+02	ug/g
18 SAMPLE	S98T002609	0 A	QMSPR-A1 SN126-A1 SOLID		N/A	<	6.194e+02	ug/g
18 SAMPLE	S98T002609	0 A	QMSPR-A1 CS133-A1 SOLID		N/A	<	8.259e+00	ug/g
18 SAMPLE	S98T002609	0 A	QMSPR-A1 CS135-A1 SOLID		N/A		3.428e+01	ug/g
18 SAMPLE	S98T002609	0 A	QMSPR-A1 CS137-A1 SOLID		N/A		6.773e+01	ug/g
19 DUP	S98T002609	0 A	QMSPR-A1 TC99-A1 SOLID		<8.26e0	<7.76e0		RPD
19 DUP	S98T002609	0 A	QMSPR-A1 SN117-A1 SOLID		<6.19e2	<5.82e2		RPD
19 DUP	S98T002609	0 A	QMSPR-A1 SN126-A1 SOLID		<6.19e2	<5.82e2		RPD
19 DUP	S98T002609	0 A	QMSPR-A1 CS133-A1 SOLID		<8.26e0	<7.76e0		RPD
19 DUP	S98T002609	0 A	QMSPR-A1 CS135-A1 SOLID		3.43e+01	2.91e+01	16.404	RPD
19 DUP	S98T002609	0 A	QMSPR-A1 CS137-A1 SOLID		6.77e+01	6.02e+01	11.728	RPD
20 CCV		0	QMSPR-QC TC99	QC	0.0205	1.90e-02	92.683	% Recovery
20 CCV		0	QMSPR-QC SN117	QC	0.0200	2.19e-02	109.500	% Recovery
20 CCV		0	QMSPR-QC CS133	QC	0.018560	1.95e-02	105.065	% Recovery
21 CCB		0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
21 CCB		0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
21 CCB		0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
21 CCB		0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
21 CCB		0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
21 CCB		0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27718

Analyst Signature

Date

Analyst Signature

Date

Saul T. Pung 03/10/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27718

Analyst: BTG Instrument: ICPMS1Book# 49857, 49857Method: LA-506-101 Rev/Mod A-OPOST SPINES - .1ml of 49857
.250ml of 10ppm Sn
(S 9015-1)

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSPR-QC	QC		
2	ICB		@MSPR-QC	QC		
3	BLNK-PREP		@MSPR-A1	SOLID		
4	SAMPLE	S98T003394 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
5	DUP	S98T003394 0 A	@MSPR-A1	SOLID		
6	SAMPLE	S98T003399 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
7	DUP	S98T003399 0 A	@MSPR-A1	SOLID		
8	SAMPLE	S98T003389 0 A	@MSPR-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
9	DUP	S98T003389 0 A	@MSPR-A1	SOLID		
10	SPK-POST	S98T003389 0 A	@MSPR-A1	SOLID		
11	CCV		@MSPR-QC	QC		
12	CCB		@MSPR-QC	QC		
13	BLNK-PREP		@MSPR-A1	SOLID		
14	SAMPLE	S98T002607 0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
15	DUP	S98T002607 0 A	@MSPR-A1	SOLID		

Data Entry Comments:

Uploaded & Validated by:
Sam M. Fung
03/10/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 27718

S Type	Sample#	R A	Test	Matrix	Group#	Project
16 SAMPLE	S98T002608	0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
17 DUP	S98T002608	0 A	@MSPR-A1	SOLID		
18 SAMPLE	S98T002609	0 A	@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
19 DUP	S98T002609	0 A	@MSPR-A1	SOLID		
20 CCV			@MSPR-QC	QC		
21 CCB			@MSPR-QC	QC		

Final page for worklist # 27718

Signature	Date
<i>[Signature]</i>	3-9-99
PREPBLK	1ml - 9ml
S98T003394	.05ml - 10ml
3394-D	.05ml - 10ml
3399	.05ml - 10ml - 1ml - 10ml
3399-D	.05ml - 10ml - 1ml - 10ml
3389	.05ml - 10ml
3389-D	.05ml - 10ml
3389-A	.05ml - 10ml
3389-A1	.05ml - 10ml
PREPBLK	1ml - 9ml
S98T002607	.05ml - 10ml
2607-D	.05ml - 10ml
2608	.05ml - 10ml - 1ml - 10ml
2608-D	.05ml - 10ml - 1ml - 10ml
2609	.05ml - 10ml - 1ml - 10ml
2609-D	.05ml - 10ml - 1ml - 10ml

Signature	Date
<i>[Signature]</i>	D.F.
10	
201	
201	
20, 201	
20, 201	
201	
201	
201	
201	
201	
10	
201	
201	
20, 201	
20, 201	
20, 201	
20, 201	

Data Entry Comments:

SCANS = 3394, 3399, 3389, PREPBLK - pb2607, 2607, 2608, 2609

LABCORE Data Entry Template for Worklist# 27746

Analyst: BIG Instrument: ICPMS1 Book# 48857, 48857Method: LA-506-101 Rev/Mod A-U

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC	QC		
2 ICB			@MSPR-QC	QC		
3 BLNK-PREP			@MSPR-A1	SOLID		
4 SAMPLE	S98T002713 0 A		@MSPR-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
5 DUP	S98T002713 0 A		@MSPR-A1	SOLID		
6 SAMPLE	S98T002846 0 A		@MSPR-A1	SOLID	98000437	BY-105 (2)
Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1						
7 DUP	S98T002846 0 A		@MSPR-A1	SOLID		
8 CCV			@MSPR-QC	QC		
9 CCB			@MSPR-QC	QC		

Final page for worklist # 27746

BIG 3-8-99
Signature Date

Signature Date

PREPBLK 1ml-9ml DF.
S98T002713 .05ml-10ml 10
2713-D .05ml-10ml 201
2846 .05ml-10ml 201
2846-D .05ml-10ml 201

Data Entry Comments:

SCANS = PREPBLK - pb 2713, 2713, 2846

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

HNF-1663 REV. 0

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/09/99 12:53 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	194000.
Stddev	1313.
%RSD	.6770

#1	195200.
#2	193700.
#3	193800.
#4	192900.
#5	192400.
#6	194700.
#7	192600.
#8	192900.
#9	195700.
#10	196000.

bjg
3-9-99
BY-105
PRV.
598700 3394
3399
3389
2713
2846
2607
2608
2609

File# PR0309A.TXT

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2869 TO 2900

Method: PRV

Sample Name: Blank

Operator:

Comment:

Run Time: 03/09/99 12:55 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.645	54.70	1.504	25.00	1.851
Stddev	.198	3.77	.253	1.12	.222
%RSD	12.01	6.889	16.80	4.490	11.97

#1	1.617	54.00	1.516	25.58	1.895
#2	1.472	56.00	1.697	26.35	1.797
#3	1.830	57.00	1.103	24.25	2.181
#4	1.440	48.50	1.465	23.52	1.566
#5	1.867	58.00	1.741	25.33	1.816

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.123	2.566	3.144	10.64
Stddev	.313	.260	.217	.43
%RSD	14.72	10.11	6.900	4.023

#1	2.552	2.401	2.932	10.69
#2	1.947	2.296	3.394	10.11
#3	2.357	2.733	3.285	11.23
#4	1.920	2.930	2.905	10.79
#5	1.842	2.472	3.204	10.37

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.875
Stddev	.109
%RSD	.55001

#1	19.785
#2	20.035
#3	19.942
#4	19.795
#5	19.819

Method: PRV

Sample Name: PRV CAL

Operator:

Comment:

Run Time: 03/09/99

12:58 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Cs[133]
Line	99/pulse	133/pulse
Units	Cts/S	Cts/S
Avg	21290.	18300.
Stddev	195.	113.
%RSD	.9168	.6163

#1	21020.	18180.
#2	21170.	18400.
#3	21340.	18400.
#4	21500.	18190.
#5	21430.	18340.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.733
Stddev	.073
%RSD	.37101

#1	19.771
#2	19.731
#3	19.618
#4	19.815
#5	19.729

Method: PRV

Sample Name: PRV1 CAL

Operator:

Comment:

Run Time: 03/09/99 13:00 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Ru[102]	Te[125]	Cs[135]	Cs[137]
Line	102/pulse	125/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	703400.	1472.	6246.	10820.
Stddev	5886.	14.	59.	193.
%RSD	.8369	.9538	.9389	1.782
#1	710900.	1477.	6343.	11130.
#2	708400.	1493.	6241.	10650.
#3	700900.	1461.	6195.	10730.
#4	697800.	1470.	6248.	10860.
#5	698900.	1458.	6204.	10720.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	65850.
Stddev	1009.
%RSD	1.532

#1	67460.
#2	65280.
#3	65370.
#4	66180.
#5	64940.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.611
Stddev	.246
%RSD	1.2565

#1	19.254
#2	19.681
#3	19.590
#4	19.590
#5	19.942

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1080.8412	1.6454	1.0000000	03/09/99 12:59:27
Ru[102]	7033.2065	54.7000	1.0000000	03/09/99 13:02:01
Te[125]	14.7044	1.5045	1.0000000	03/09/99 13:02:01
Sn[117]	946.6558	25.0039	1.0000000	03/09/99 12:57:14
Sn[126]	947.8586	1.8512	1.0000000	03/09/99 12:57:14
Cs[133]	915.0640	2.1235	1.0000000	03/09/99 12:59:27
Cs[135]	945.9816	2.5664	1.0000000	03/09/99 13:02:01
Cs[137]	965.4728	3.1438	1.0000000	03/09/99 13:02:01
Ba[138]	658.3510	10.6369	1.0000000	03/09/99 13:02:01

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV-SN CAL

Operator:

Comment:

Run Time: 03/09/99 13:13 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Sn[117]	Sn[126]
Line	117/pulse	126/pulse
Units	Cts/S	Cts/S
Avg	16500.	16500.
Stddev	428.	428.
%RSD	2.594	2.594

#1	17150.	17150.
#2	16150.	16150.
#3	16510.	16510.
#4	16080.	16080.
#5	16640.	16640.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.257
Stddev	.334
%RSD	1.7336

#1	18.691
#2	19.388
#3	19.260
#4	19.387
#5	19.559

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1080.8412	1.6454	1.0000000	03/09/99 12:59:27
Ru[102]	7033.2065	54.7000	1.0000000	03/09/99 13:02:01
Te[125]	14.7044	1.5045	1.0000000	03/09/99 13:02:01
Sn[117]	856.0545	25.0039	1.0000000	03/09/99 13:14:41
Sn[126]	857.2572	1.8512	1.0000000	03/09/99 13:14:41
Cs[133]	915.0640	2.1235	1.0000000	03/09/99 12:59:27
Cs[135]	945.9816	2.5664	1.0000000	03/09/99 13:02:01
Cs[137]	965.4728	3.1438	1.0000000	03/09/99 13:02:01
Ba[138]	658.3510	10.6369	1.0000000	03/09/99 13:02:01

Method: PRV

Sample Name: icv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 13:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.23	.0005	.0124	1.418	.0003
Stddev	.49	.0011	.0195	.173	.0002
%RSD	2.564	236.9	157.1	12.22	71.03
#1	20.11	-.0007	-.0065	1.187	.0004
#2	19.14	.0003	.0161	1.302	.0001
#3	19.00	.0020	.0092	1.457	.0005
#4	19.01	-.0004	.0436	1.530	.0001
#5	18.92	.0012	-.0003	1.616	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.83	.0069	.0123	.1147
Stddev	.42	.0006	.0009	.0032
%RSD	2.105	8.250	7.566	2.802
#1	20.32	.0078	.0131	.1181
#2	19.17	.0065	.0115	.1153
#3	19.86	.0071	.0114	.1130
#4	19.97	.0064	.0119	.1101
#5	19.81	.0069	.0134	.1169

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.030
Stddev	.191
%RSD	1.0034

#1	18.819
#2	19.245
#3	19.205
#4	18.876
#5	19.005

Method: PRV

Sample Name: icv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 13:25 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1524	.0006	.0301	21.39	.0004
Stddev	.1494	.0010	.0091	.44	.0004
%RSD	98.04	160.1	30.32	2.044	104.6

#1	.4031	.0014	.0344	20.68	.0004
#2	.1731	-.0003	.0275	21.59	.0009
#3	.0919	-.0005	.0298	21.85	.0006
#4	.0563	.0009	.0419	21.42	.0003
#5	.0377	.0016	.0171	21.39	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0127	.0282	.0444	.4179
Stddev	.0031	.0018	.0013	.0213
%RSD	24.62	6.545	2.891	5.092

#1	.0177	.0267	.0455	.4072
#2	.0129	.0275	.0426	.4324
#3	.0123	.0301	.0453	.4473
#4	.0110	.0304	.0451	.3951
#5	.0095	.0266	.0434	.4075

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.733
Stddev	.596
%RSD	3.1788

#1	19.643
#2	18.864
#3	18.017
#4	18.630
#5	18.511

Method: PRV

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0010	.0063	-.0002	.0006
Stddev	.0002	.0010	.0265	.0007	.0005
%RSD	56.92	103.6	419.8	402.9	81.29

#1	.0003	.0010	-.0039	.0008	.0011
#2	.0001	-.0002	.0024	-.0005	-.0002
#3	.0005	.0020	.0213	.0002	.0005
#4	.0001	.0001	-.0292	-.0010	.0008
#5	.0004	.0019	.0411	-.0004	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0007	-.0002	-.0002	-.0002
Stddev	.0003	.0002	.0003	.0009
%RSD	41.85	68.20	173.5	533.4

#1	.0004	-.0002	-.0003	-.0005
#2	.0006	.0000	.0000	.0003
#3	.0008	-.0004	.0002	-.0002
#4	.0006	-.0003	-.0004	.0009
#5	.0012	-.0001	-.0003	-.0015

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.975
Stddev	.092
%RSD	.48244

#1	19.011
#2	18.827
#3	18.987
#4	19.076
#5	18.972

HNF-1663 REV. 0

Method: PRV

Sample Name: prepbklk

Operator: bjg

Comment: 10 df

Run Time: 03/09/99 13:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0147	.0002	.0007	.4405	.0002
Stddev	.0076	.0008	.0140	.0028	.0002
%RSD	51.53	438.5	2086.	.6415	121.8

#1	.0257	.0006	.0000	.4376	.0005
#2	.0186	-.0004	.0205	.4402	.0000
#3	.0128	.0008	-.0071	.4429	.0000
#4	.0091	-.0010	.0065	.4440	.0002
#5	.0071	.0009	-.0165	.4381	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0011	.0043	.0068	.0654
Stddev	.0003	.0003	.0007	.0024
%RSD	24.31	7.402	9.963	3.635

#1	.0010	.0048	.0068	.0643
#2	.0016	.0043	.0073	.0690
#3	.0010	.0040	.0075	.0657
#4	.0009	.0041	.0057	.0626
#5	.0011	.0044	.0068	.0653

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	20.274
Stddev	.062
%RSD	.30428

#1	20.275
#2	20.208
#3	20.352
#4	20.317
#5	20.220

Method: PRV

Sample Name: s98t003394

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 13:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0068	.0064	.0080	.0310	.0004
Stddev	.0002	.0011	.0124	.0013	.0003
%RSD	3.585	17.96	155.8	4.198	88.84

#1	.0066	.0067	.0167	.0310	.0006
#2	.0069	.0057	.0149	.0324	-.0001
#3	.0072	.0061	.0101	.0316	.0002
#4	.0067	.0081	.0119	.0289	.0005
#5	.0067	.0052	-.0138	.0309	.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0072	.0035	.0112	.0200
Stddev	.0006	.0005	.0007	.0015
%RSD	8.172	15.19	6.191	7.499

#1	.0066	.0034	.0100	.0214
#2	.0075	.0042	.0116	.0209
#3	.0081	.0037	.0113	.0198
#4	.0072	.0031	.0118	.0206
#5	.0068	.0029	.0112	.0176

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.861
Stddev	.319
%RSD	1.6065

#1	20.000
#2	19.430
#3	19.658
#4	20.248
#5	19.970

Method: PRV   HNF-1663 REV. 0
 Comment: 201 df   Sample Name: s98t003394_d   Operator: bjg
 Run Time: 03/09/99 13:54 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0075	.0075	.0012	.0335	.0002
Stddev	.0006	.0018	.0127	.0026	.0003
%RSD	7.536	24.40	1049.	7.903	123.7

#1	.0072	.0064	-.0148	.0324	.0000
#2	.0082	.0069	.0143	.0355	.0000
#3	.0067	.0073	.0085	.0293	.0002
#4	.0079	.0108	-.0097	.0346	.0007
#5	.0076	.0064	.0076	.0355	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0081	.0030	.0065	.0101
Stddev	.0006	.0004	.0007	.0015
%RSD	7.754	13.66	10.45	15.09

#1	.0079	.0026	.0065	.0107
#2	.0086	.0035	.0062	.0085
#3	.0088	.0033	.0073	.0094
#4	.0080	.0031	.0056	.0124
#5	.0072	.0026	.0069	.0094

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.587
Stddev	.243
%RSD	1.2388

#1	19.817
#2	19.242
#3	19.631
#4	19.449
#5	19.793

Method: PRV

Sample Name: s98t003399

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 13:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	.0026	-.0075	-.0060	.0002
Stddev	.0002	.0007	.0139	.0003	.0002
%RSD	45.40	28.90	186.3	5.306	134.7

#1	.0001	.0032	-.0187	-.0057	-.0002
#2	.0004	.0032	-.0181	-.0060	.0001
#3	.0004	.0017	.0137	-.0061	.0003
#4	.0004	.0018	-.0006	-.0058	.0003
#5	.0006	.0029	-.0135	-.0066	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0013	.0069	.0139	.1131
Stddev	.0003	.0007	.0008	.0032
%RSD	26.13	10.72	5.862	2.793

#1	.0011	.0079	.0134	.1107
#2	.0017	.0067	.0136	.1132
#3	.0011	.0064	.0153	.1096
#4	.0010	.0061	.0132	.1175
#5	.0016	.0073	.0138	.1147

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.454
Stddev	.392
%RSD	2.0131

#1	19.935
#2	19.790
#3	19.348
#4	19.051
#5	19.146

Method: PRV

Sample Name: s98t003399_d

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 14:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	.0028	.0002	-.0084	.0002
Stddev	.0002	.0010	.0137	.0003	.0002
%RSD	29.12	36.33	6278.	3.963	88.40

#1	.0007	.0022	.0039	-.0084	.0002
#2	.0006	.0039	.0022	-.0079	.0003
#3	.0004	.0037	.0149	-.0085	.0005
#4	.0005	.0027	.0025	-.0083	.0000
#5	.0008	.0015	-.0224	-.0089	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0012	.0079	.0161	.1304
Stddev	.0001	.0004	.0012	.0048
%RSD	11.96	4.963	7.173	3.715

#1	.0013	.0085	.0164	.1236
#2	.0013	.0080	.0180	.1335
#3	.0012	.0075	.0151	.1360
#4	.0013	.0081	.0156	.1280
#5	.0010	.0077	.0153	.1311

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.045
Stddev	.167
%RSD	.87531

#1	19.209
#2	18.872
#3	18.858
#4	19.134
#5	19.153

Method: PRV   Sample Name: s98t003389   Operator: bjg
Comment: 20ldf
Run Time: 03/09/99 14:03 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0113	.0098	-.0043	.0113	.0006
Stddev	.0007	.0009	.0160	.0010	.0003
%RSD	5.936	9.236	369.1	8.771	52.83

#1	.0116	.0104	-.0211	.0113	.0004
#2	.0116	.0083	-.0213	.0110	.0007
#3	.0121	.0106	.0026	.0110	.0007
#4	.0105	.0096	.0043	.0129	.0001
#5	.0107	.0101	.0138	.0102	.0009

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0128	.0060	.0138	.0229
Stddev	.0005	.0005	.0010	.0022
%RSD	3.768	7.588	7.536	9.519

#1	.0124	.0064	.0128	.0202
#2	.0135	.0055	.0154	.0254
#3	.0125	.0062	.0143	.0248
#4	.0132	.0063	.0133	.0214
#5	.0126	.0055	.0132	.0228

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.274
Stddev	.524
%RSD	2.7171

#1	19.689
#2	18.470
#3	19.115
#4	19.775
#5	19.321

Method: PRV

Sample Name: s98t003389_d

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0117	.0097	.0029	.0543	.0002
Stddev	.0006	.0016	.0159	.0010	.0003
%RSD	5.109	16.71	550.1	1.875	111.5

#1	.0118	.0103	-.0101	.0559	.0001
#2	.0117	.0073	.0230	.0541	.0003
#3	.0107	.0091	.0052	.0544	-.0001
#4	.0124	.0105	-.0158	.0530	.0007
#5	.0118	.0116	.0121	.0544	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0136	.0057	.0148	.0240
Stddev	.0008	.0003	.0007	.0009
%RSD	5.741	5.720	4.537	3.860

#1	.0127	.0053	.0156	.0243
#2	.0137	.0060	.0153	.0252
#3	.0136	.0057	.0139	.0227
#4	.0132	.0054	.0147	.0235
#5	.0148	.0060	.0145	.0242

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.784
Stddev	.342
%RSD	1.7302

#1	19.903
#2	19.802
#3	20.247
#4	19.656
#5	19.311

Method: PRV

Sample Name: s98t003389_a

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:10 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	20.93	.0090	-.0142	3.572	.0002
Stddev	.55	.0014	.0146	.081	.0002
%RSD	2.639	16.17	103.3	2.254	121.7

#1	20.16	.0088	-.0170	3.449	-.0001
#2	21.13	.0114	-.0230	3.611	.0001
#3	21.63	.0082	-.0010	3.645	.0005
#4	20.66	.0076	.0024	3.532	.0001
#5	21.05	.0087	-.0322	3.622	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	20.46	.0072	.0160	.0397
Stddev	.59	.0006	.0007	.0015
%RSD	2.860	8.241	4.290	3.681

#1	19.58	.0070	.0154	.0389
#2	20.48	.0073	.0162	.0385
#3	21.22	.0072	.0170	.0419
#4	20.55	.0081	.0161	.0405
#5	20.48	.0065	.0153	.0386

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.612
Stddev	.444
%RSD	2.2652

#1	20.317
#2	19.711
#3	19.139
#4	19.491
#5	19.402

HNF-1663 REV.0

Method: PRV

Sample Name: s98t003389_a1

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0215	.0101	.0070	18.92	.0002
Stddev	.0070	.0008	.0159	.21	.0003
%RSD	32.71	8.052	227.8	1.120	160.2
#1	.0333	.0096	.0225	19.28	.0004
#2	.0224	.0100	.0198	18.71	-.0003
#3	.0179	.0100	.0111	18.93	.0001
#4	.0168	.0114	-.0035	18.83	.0005
#5	.0169	.0093	-.0150	18.89	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0218	.0082	.0157	.0275
Stddev	.0041	.0007	.0007	.0009
%RSD	18.67	8.355	4.471	3.251
#1	.0288	.0089	.0157	.0283
#2	.0220	.0077	.0150	.0264
#3	.0201	.0076	.0152	.0267
#4	.0186	.0090	.0168	.0276
#5	.0197	.0080	.0158	.0284

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.967
Stddev	.263
%RSD	1.3169

#1	19.614
#2	20.321
#3	20.084
#4	19.966
#5	19.852

Method: PRV

Sample Name: ccv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 14:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0065	-.0007	.0136	21.18	.0003
Stddev	.0039	.0006	.0131	1.25	.0004
%RSD	60.62	87.09	95.82	5.907	118.3

#1	.0122	-.0015	-.0029	19.04	.0002
#2	.0088	-.0004	.0131	21.40	.0010
#3	.0045	-.0005	.0151	21.85	.0003
#4	.0033	.0000	.0095	21.36	.0000
#5	.0035	-.0010	.0333	22.26	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0097	.0272	.0443	.4151
Stddev	.0007	.0014	.0022	.0290
%RSD	7.365	5.321	5.001	6.982

#1	.0095	.0259	.0407	.3654
#2	.0104	.0269	.0454	.4223
#3	.0100	.0296	.0461	.4389
#4	.0085	.0275	.0436	.4174
#5	.0100	.0263	.0455	.4313

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.102
Stddev	1.391
%RSD	7.2792

#1	21.551
#2	18.558
#3	18.244
#4	18.860
#5	18.298

Method: PRV Sample Name: ccv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 14:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.35	.0021	-.0015	1.515	.0002
Stddev	.29	.0013	.0083	.095	.0004
%RSD	1.500	61.91	551.5	6.268	226.0
#1	19.86	.0037	-.0133	1.439	.0001
#2	19.14	.0025	.0022	1.406	.0003
#3	19.27	.0021	-.0069	1.515	.0003
#4	19.26	.0020	.0039	1.586	-.0004
#5	19.20	.0001	.0065	1.630	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.65	.0075	.0124	.1165
Stddev	.23	.0003	.0008	.0028
%RSD	1.159	4.390	6.098	2.402
#1	20.01	.0075	.0130	.1197
#2	19.64	.0080	.0129	.1137
#3	19.43	.0076	.0114	.1151
#4	19.68	.0072	.0117	.1147
#5	19.48	.0072	.0129	.1193

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.697
Stddev	.198
%RSD	1.0057

#1	19.480
#2	19.837
#3	19.950
#4	19.535
#5	19.683

Method: PRV Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 14:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0018	.0007	.0091	.0001	.0001
Stddev	.0003	.0004	.0135	.0010	.0003
%RSD	15.81	53.97	147.8	1500.	243.9

#1	.0014	.0003	.0121	.0013	.0005
#2	.0022	.0008	.0182	-.0003	-.0004
#3	.0017	.0011	-.0057	-.0010	.0002
#4	.0017	.0011	-.0038	-.0007	.0002
#5	.0019	.0003	.0250	.0010	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0001	.0000	.0005
Stddev	.0003	.0002	.000	.0005
%RSD	46.51	233.3	588.0	102.5

#1	.0006	.0000	-.0001	-.0002
#2	.0006	.0001	-.0001	.0005
#3	.0003	.0000	.0004	.0003
#4	.0010	-.0004	-.0003	.0012
#5	.0005	-.0001	-.0002	.0008

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.178
Stddev	.460
%RSD	2.3987

#1	18.430
#2	19.184
#3	19.365
#4	19.677
#5	19.233

Method: PRV Sample Name: prepbk

Operator: bjg

Comment: 10 df

Run Time: 03/09/99 14:51 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Te[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0048	.0006	-.0109	.3460	.0000
Stddev	.0016	.0005	.0219	.0044	.000
%RSD	32.92	74.78	201.1	1.259	374.7

#1	.0069	.0005	-.0373	.3447	-.0001
#2	.0060	.0011	-.0202	.3483	-.0001
#3	.0039	.0010	.0181	.3518	.0001
#4	.0033	-.0001	-.0191	.3450	.0002
#5	.0038	.0008	.0041	.3402	-.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0013	.0050	.0085	.0813
Stddev	.0003	.0005	.0004	.0022
%RSD	20.54	9.333	4.289	2.766

#1	.0014	.0050	.0090	.0847
#2	.0016	.0058	.0088	.0823
#3	.0014	.0051	.0084	.0806
#4	.0012	.0046	.0081	.0791
#5	.0009	.0048	.0083	.0799

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	20.384
Stddev	.065
%RSD	.31966

#1	20.406
#2	20.301
#3	20.330
#4	20.429
#5	20.452

Method: PRV

Sample Name: s98t002607

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0082	.0141	-.0074	.0138	.0001
Stddev	.0005	.0018	.0118	.0026	.0001
%RSD	6.291	12.86	158.4	19.02	83.85

#1	.0083	.0135	.0055	.0164	.0002
#2	.0085	.0155	-.0016	.0148	.0002
#3	.0075	.0126	-.0253	.0104	.0001
#4	.0079	.0165	-.0039	.0157	.0001
#5	.0088	.0123	-.0120	.0117	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0077	.0030	.0083	.0085
Stddev	.0003	.0002	.0010	.0006
%RSD	3.740	7.181	11.73	7.486

#1	.0075	.0027	.0074	.0082
#2	.0079	.0028	.0074	.0086
#3	.0075	.0030	.0091	.0076
#4	.0081	.0030	.0095	.0087
#5	.0076	.0032	.0079	.0093

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.853
Stddev	.310
%RSD	1.5635

#1	20.182
#2	19.911
#3	19.879
#4	19.340
#5	19.955

Method: PRV ²⁰¹ Sample Name: s98t002607_d Operator: bjg
 Comment: ~~20301~~ df
 Run Time: 03/09/99 15:00 Type: Unk Mode: CONC Corr.Fact: 1.000000

SP 3/10/99

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0068	.0110	-.0032	.0120	.0000
Stddev	.0005	.0020	.0138	.0010	.0002
%RSD	7.843	17.75	432.4	8.021	463.8

#1	.0071	.0140	-.0053	.0109	-.0001
#2	.0070	.0105	-.0011	.0133	.0000
#3	.0059	.0114	-.0244	.0126	.0000
#4	.0072	.0106	.0016	.0117	.0003
#5	.0068	.0086	.0133	.0113	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0067	.0025	.0073	.0066
Stddev	.0004	.0004	.0006	.0009
%RSD	6.199	14.77	8.013	13.63

#1	.0067	.0026	.0066	.0077
#2	.0073	.0028	.0075	.0070
#3	.0067	.0029	.0081	.0058
#4	.0061	.0020	.0071	.0056
#5	.0068	.0022	.0070	.0071

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	20.092
Stddev	.205
%RSD	1.0186

#1	19.976
#2	19.818
#3	20.087
#4	20.295
#5	20.286

HNF-1663 REV. 0

Method: PRV

Sample Name: s98t002608

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 15:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0007	.0041	.0085	-.0070	.0002
Stddev	.0003	.0010	.0107	.0007	.0002
%RSD	50.92	25.12	126.0	9.439	92.75

#1	.0010	.0044	.0195	-.0068	.0004
#2	.0009	.0048	-.0073	-.0063	.0003
#3	.0007	.0028	.0063	-.0066	.0003
#4	.0006	.0053	.0174	-.0080	-.0001
#5	.0001	.0033	.0066	-.0073	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0012	.0073	.0157	.1251
Stddev	.0004	.0003	.0009	.0028
%RSD	30.42	4.412	5.715	2.215

#1	.0012	.0071	.0150	.1267
#2	.0017	.0071	.0170	.1231
#3	.0008	.0076	.0159	.1217
#4	.0016	.0076	.0147	.1286
#5	.0010	.0069	.0157	.1256

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.397
Stddev	.244
%RSD	1.2566

#1	19.535
#2	19.688
#3	19.409
#4	19.314
#5	19.041

Method: PRV

Sample Name: s98t002608_d

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 15:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0006	.0038	.0076	-.0083	.0002
Stddev	.0001	.0003	.0084	.0007	.0003
%RSD	24.28	7.537	110.2	8.845	118.9

#1	.0008	.0038	.0202	-.0076	-.0001
#2	.0005	.0040	.0106	-.0092	.0004
#3	.0005	.0041	.0068	-.0089	.0005
#4	.0007	.0034	-.0008	-.0083	.0004
#5	.0004	.0037	.0012	-.0075	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0010	.0068	.0140	.1118
Stddev	.0001	.0005	.0007	.0016
%RSD	14.55	7.997	5.006	1.458

#1	.0007	.0075	.0137	.1126
#2	.0010	.0071	.0150	.1120
#3	.0011	.0063	.0144	.1122
#4	.0010	.0064	.0134	.1133
#5	.0009	.0064	.0133	.1090

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.241
Stddev	.211
%RSD	1.0983

#1	19.425
#2	18.977
#3	19.324
#4	19.425
#5	19.055

Method: PRV

Sample Name: s98t002609

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 15:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc [99]	Ru [102]	Te [125]	Sn [117]	Sn [126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0005	.0032	.0039	-.0092	.0003
Stddev	.0001	.0013	.0131	.0018	.0002
%RSD	31.78	39.66	339.8	19.78	59.84

#1	.0004	.0014	.0017	-.0075	.0001
#2	.0003	.0046	-.0178	-.0107	.0002
#3	.0004	.0032	.0132	-.0076	.0002
#4	.0007	.0027	.0143	-.0087	.0005
#5	.0004	.0042	.0079	-.0116	.0004

Elem	Cs [133]	Cs [135]	Cs [137]	Ba [138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0010	.0083	.0164	.1370
Stddev	.0002	.0009	.0007	.0082
%RSD	25.09	11.34	4.545	5.953

#1	.0009	.0079	.0153	.1360
#2	.0008	.0091	.0165	.1343
#3	.0014	.0088	.0167	.1489
#4	.0010	.0086	.0173	.1395
#5	.0008	.0068	.0164	.1265

Int. Std.	In [115]
Line	115/pulse
Units	Cts/S
Avg	18.487
Stddev	.667
%RSD	3.6097

#1	18.961
#2	18.911
#3	17.370
#4	18.367
#5	18.825

Method: PRV

Sample Name: s98t002609_d

Operator: bjg

Comment: 20301 df

Run Time: 03/09/99 15:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0004	.0042	.0061	-.0098	.0002
Stddev	.0002	.0011	.0130	.0008	.0001
%RSD	63.45	25.82	214.2	7.779	56.00

#1	.0001	.0025	.0076	-.0106	.0001
#2	.0003	.0041	-.0012	-.0091	.0001
#3	.0008	.0055	.0177	-.0095	.0002
#4	.0003	.0042	-.0122	-.0107	.0003
#5	.0003	.0045	.0184	-.0092	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0007	.0075	.0155	.1262
Stddev	.0003	.0008	.0011	.0046
%RSD	50.64	9.975	6.853	3.671

#1	.0008	.0086	.0153	.1210
#2	.0006	.0073	.0144	.1219
#3	.0010	.0078	.0169	.1280
#4	.0001	.0065	.0145	.1285
#5	.0009	.0075	.0162	.1318

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.950
Stddev	.203
%RSD	1.0711

#1	18.864
#2	19.169
#3	18.701
#4	18.866
#5	19.151

HNF-1663 REV. 0

Method: PRV Sample Name: ccv

Operator: bjj

Comment: msprv lcs

Run Time: 03/09/99 15:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0016	-.0004	.0198	21.86	.0000
Stddev	.0007	.0009	.0098	1.15	.000
%RSD	42.70	246.0	49.38	5.273	3962.

#1	.0024	-.0010	.0188	22.36	.0000
#2	.0015	-.0012	.0225	21.44	-.0004
#3	.0020	-.0007	.0347	21.70	-.0003
#4	.0014	.0009	.0092	23.46	.0007
#5	.0006	.0003	.0137	20.34	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0073	.0282	.0447	.4165
Stddev	.0012	.0013	.0024	.0073
%RSD	16.55	4.578	5.355	1.752

#1	.0079	.0275	.0463	.4204
#2	.0065	.0271	.0430	.4113
#3	.0059	.0290	.0445	.4191
#4	.0090	.0301	.0477	.4248
#5	.0074	.0273	.0418	.4067

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.205
Stddev	.591
%RSD	3.2485

#1	17.961
#2	18.800
#3	18.115
#4	17.388
#5	18.762

Method: PRV Sample Name: ccv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 15:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.03	.0032	.0035	1.443	.0001
Stddev	.88	.0003	.0127	.158	.0002
%RSD	4.640	9.593	365.7	10.94	330.8

#1	17.46	.0036	.0004	1.195	-.0003
#2	19.40	.0030	-.0084	1.392	.0002
#3	19.58	.0035	.0250	1.486	.0002
#4	19.41	.0032	.0028	1.544	.0000
#5	19.32	.0029	-.0024	1.596	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.53	.0069	.0121	.1130
Stddev	.74	.0006	.0008	.0044
%RSD	3.766	9.131	6.314	3.885

#1	18.25	.0060	.0112	.1053
#2	20.02	.0072	.0130	.1155
#3	19.89	.0076	.0122	.1147
#4	19.55	.0073	.0114	.1134
#5	19.93	.0066	.0126	.1160

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.549
Stddev	.746
%RSD	3.8142

#1	20.863
#2	19.191
#3	19.233
#4	19.408
#5	19.049

Method: PRV

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 15:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0015	.0010	-.0009	-.0009	.0001
Stddev	.0001	.0015	.0064	.0016	.0004
%RSD	4.753	157.1	721.0	177.1	368.0

#1	.0014	-.0001	.0061	-.0024	.0001
#2	.0015	-.0007	-.0018	-.0012	.0002
#3	.0016	.0017	.0047	.0012	-.0005
#4	.0015	.0031	-.0041	-.0025	.0003
#5	.0015	.0008	-.0093	.0004	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0001	-.0001	-.0005
Stddev	.0005	.0002	.0003	.0012
%RSD	72.24	164.3	364.9	271.8

#1	.0002	.0001	.0001	.0011
#2	.0013	-.0001	-.0002	.0002
#3	.0009	.0000	-.0005	-.0016
#4	.0006	-.0002	-.0001	.0000
#5	.0002	-.0004	.0003	-.0018

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.583
Stddev	.289
%RSD	1.5544

#1	18.184
#2	18.948
#3	18.431
#4	18.698
#5	18.653

LABCORE Completed Worklist Report for Worklist# 27741

Analyst: kjt Instrument: ICPMS1

Book#: 11B59

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0		EMSACT TH232	QC	1.00e-2	1.02e-02	102.000 %	Recovery
1 ICV	0		EMSACT NP237	QC	9.88e-3	1.13e-02	114.372 %	Recovery
1 ICV	0		EMSACT PU239	QC	1.004e-2	9.89e-03	98.506 %	Recovery
1 ICV	0		EMSACT 241	QC	1.026e-2	1.05e-02	102.339 %	Recovery
2 ICB	0		EMSAC-QC PA231	QC	1	<3.30e-6		ug/mL
2 ICB	0		EMSAC-QC TH229	QC	1	<4.34e-6		ug/mL
2 ICB	0		EMSAC-QC TH230	QC	1	<3.21e-6		ug/mL
2 ICB	0		EMSAC-QC TH232	QC	1	<2.80e-6		ug/mL
2 ICB	0		EMSAC-QC NP237	QC	1	<1.48e-6		ug/mL
2 ICB	0		EMSAC-QC PU239	QC	1	<3.23e-6		ug/mL
2 ICB	0		EMSAC-QC PU240	QC	1	<2.18e-6		ug/mL
2 ICB	0		EMSAC-QC PU242	QC	1	<1.54e-6		ug/mL
2 ICB	0		EMSAC-QC 241	QC	1	<1.79e-6		ug/mL
2 ICB	0		EMSAC-QC 243	QC	1	<2.60e-6		ug/mL
2 ICB	0		EMSAC-QC 244	QC	1	<3.82e-6		ug/mL
3 BLNK-PREP	0		EMSACTAL PA231-A1	SOLID	1	<3.63e-5		ug/g
3 BLNK-PREP	0		EMSACTAL TH229-A1	SOLID	1	<4.77e-5		ug/g
3 BLNK-PREP	0		EMSACTAL TH230-A1	SOLID	1	<3.53e-5		ug/g
3 BLNK-PREP	0		EMSACTAL TH232-A1	SOLID	1	<3.08e-5		ug/g
3 BLNK-PREP	0		EMSACTAL NP237-A1	SOLID	1	<1.63e-5		ug/g
3 BLNK-PREP	0		EMSACTAL PU239-A1	SOLID	1	<3.56e-5		ug/g
3 BLNK-PREP	0		EMSACTAL PU240-A1	SOLID	1	<2.40e-5		ug/g
3 BLNK-PREP	0		EMSACTAL PU242-A1	SOLID	1	<1.69e-5		ug/g
3 BLNK-PREP	0		EMSACTAL 241-A1	SOLID	1	<1.97e-5		ug/g
3 BLNK-PREP	0		EMSACTAL 243-A1	SOLID	1	<2.86e-5		ug/g
3 BLNK-PREP	0		EMSACTAL 244-A1	SOLID	1	<4.20e-5		ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL PA231-A1	SOLID	N/A	< 5.647e+00	5.647	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL TH229-A1	SOLID	N/A	< 7.428e+00	7.428	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL TH230-A1	SOLID	N/A	< 5.498e+00	5.498	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL TH232-A1	SOLID	N/A	< 4.784e+00	4.784	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL NP237-A1	SOLID	N/A	< 2.529e+00	2.529	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL PU239-A1	SOLID	N/A	< 5.532e+00	5.532	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL PU240-A1	SOLID	N/A	< 3.728e+00	3.727	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL PU242-A1	SOLID	N/A	< 2.631e+00	2.631	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL 241-A1	SOLID	N/A	< 3.063e+00	3.063	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL 243-A1	SOLID	N/A	< 4.446e+00	4.446	ug/g
4 SAMPLE	S98T002847	0 A	EMSACTAL 244-A1	SOLID	N/A	< 6.531e+00	6.531	ug/g
5 DUP	S98T002847	0 A	EMSACTAL PA231-A1	SOLID	<5.65e0	<5.13e0		RPD
5 DUP	S98T002847	0 A	EMSACTAL TH229-A1	SOLID	<7.43e0	<6.75e0		RPD
5 DUP	S98T002847	0 A	EMSACTAL TH230-A1	SOLID	<5.50e0	<5.00e0		RPD
5 DUP	S98T002847	0 A	EMSACTAL TH232-A1	SOLID	<4.78e0	<4.35e0		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27741

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 DUP	S98T002847	0 A	OMSACT1	NP237-A1 SOLID	<2.53e0	<2.30e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	PU239-A1 SOLID	<5.53e0	<5.03e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	PU240-A1 SOLID	<3.73e0	<3.39e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	PU242-A1 SOLID	<2.63e0	<2.39e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	241-A1 SOLID	<3.06e0	<2.78e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	243-A1 SOLID	<4.45e0	<4.04e0		RPD
5 DUP	S98T002847	0 A	OMSACT1	244-A1 SOLID	<6.53e0	<5.93e0		RPD
6 SAMPLE	S98T002926	0 A	OMSACT1	PA231-A1 SOLID	N/A	1.313e-01	0.131	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	TH229-A1 SOLID	N/A	1.728e-01	0.173	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	TH230-A1 SOLID	N/A	1.279e-01	0.128	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	TH232-A1 SOLID	N/A	9.817e+00	0.111	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	NP237-A1 SOLID	N/A	5.882e-02	5.88e-002	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	PU239-A1 SOLID	N/A	1.287e-01	0.129	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	PU240-A1 SOLID	N/A	8.670e-02	8.67e-002	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	PU242-A1 SOLID	N/A	6.120e-02	6.12e-002	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	241-A1 SOLID	N/A	7.125e-02	7.12e-002	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	243-A1 SOLID	N/A	1.034e-01	0.103	ug/g
6 SAMPLE	S98T002926	0 A	OMSACT1	244-A1 SOLID	N/A	1.519e-01	0.152	ug/g
7 DUP	S98T002926	0 A	OMSACT1	PA231-A1 SOLID	<1.31e-1	<1.33e-1		RPD
7 DUP	S98T002926	0 A	OMSACT1	TH229-A1 SOLID	<1.73e-1	<1.75e-1		RPD
7 DUP	S98T002926	0 A	OMSACT1	TH230-A1 SOLID	<1.28e-1	<1.29e-1		RPD
7 DUP	S98T002926	0 A	OMSACT1	TH232-A1 SOLID	9.82e+00	2.48e+00	119.350	RPD
7 DUP	S98T002926	0 A	OMSACT1	NP237-A1 SOLID	<5.88e-2	<5.95e-2		RPD
7 DUP	S98T002926	0 A	OMSACT1	PU239-A1 SOLID	<1.29e-1	<1.30e-1		RPD
7 DUP	S98T002926	0 A	OMSACT1	PU240-A1 SOLID	<8.67e-2	<8.77e-2		RPD
7 DUP	S98T002926	0 A	OMSACT1	PU242-A1 SOLID	<6.12e-2	<6.19e-2		RPD
7 DUP	S98T002926	0 A	OMSACT1	241-A1 SOLID	<7.12e-2	<7.21e-2		RPD
7 DUP	S98T002926	0 A	OMSACT1	243-A1 SOLID	<1.03e-1	<1.05e-1		RPD
7 DUP	S98T002926	0 A	OMSACT1	244-A1 SOLID	<1.52e-1	<1.54e-1		RPD
8 SAMPLE	S98T002927	0 A	OMSACT1	PA231-A1 SOLID	N/A	1.300e-01	0.130	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	TH229-A1 SOLID	N/A	1.709e-01	0.171	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	TH230-A1 SOLID	N/A	1.265e-01	0.127	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	TH232-A1 SOLID	N/A	5.867e-01	0.110	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	NP237-A1 SOLID	N/A	5.820e-02	5.82e-002	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	PU239-A1 SOLID	N/A	1.273e-01	0.127	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	PU240-A1 SOLID	N/A	8.579e-02	8.58e-002	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	PU242-A1 SOLID	N/A	6.055e-02	6.05e-002	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	241-A1 SOLID	N/A	7.050e-02	7.05e-002	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	243-A1 SOLID	N/A	1.023e-01	0.102	ug/g
8 SAMPLE	S98T002927	0 A	OMSACT1	244-A1 SOLID	N/A	1.503e-01	0.150	ug/g
9 DUP	S98T002927	0 A	OMSACT1	PA231-A1 SOLID	<1.30e-1	<1.26e-1		RPD
9 DUP	S98T002927	0 A	OMSACT1	TH229-A1 SOLID	<1.71e-1	<1.66e-1		RPD
9 DUP	S98T002927	0 A	OMSACT1	TH230-A1 SOLID	<1.27e-1	<1.23e-1		RPD
9 DUP	S98T002927	0 A	OMSACT1	TH232-A1 SOLID	5.87e-01	5.69e-01	3.114	RPD
9 DUP	S98T002927	0 A	OMSACT1	NP237-A1 SOLID	<5.82e-2	<5.66e-2		RPD
9 DUP	S98T002927	0 A	OMSACT1	PU239-A1 SOLID	<1.27e-1	<1.24e-1		RPD
9 DUP	S98T002927	0 A	OMSACT1	PU240-A1 SOLID	<8.58e-2	<8.34e-2		RPD
9 DUP	S98T002927	0 A	OMSACT1	PU242-A1 SOLID	<6.05e-2	<5.89e-2		RPD
9 DUP	S98T002927	0 A	OMSACT1	241-A1 SOLID	<7.05e-2	<6.86e-2		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27741

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
9 DUP	S98T002927	0 A	OMSACTA1 243-A1	SOLID	<1.02e-1	<9.95e-2		RPD
9 DUP	S98T002927	0 A	OMSACTA1 244-A1	SOLID	<1.50e-1	<1.46e-1		RPD
10 SPK-POST	S98T002927	0 A	OMSACTA1 TH232-A1	SOLID	0.0200	1.98e-02	99.000 % Recovery	
10 SPK-POST	S98T002927	0 A	OMSACTA1 NP237-A1	SOLID	0.019760	2.25e-02	113.866 % Recovery	
10 SPK-POST	S98T002927	0 A	OMSACTA1 PU239-A1	SOLID	0.020080	1.91e-02	95.120 % Recovery	
10 SPK-POST	S98T002927	0 A	OMSACTA1 241-A1	SOLID	0.020520	2.07e-02	100.877 % Recovery	
11 CCV	0	0	OMSACT TH232	QC	1.00e-2	1.05e-02	105.000 % Recovery	
11 CCV	0	0	OMSACT NP237	QC	9.88e-3	1.18e-02	119.433 % Recovery	
11 CCV	0	0	OMSACT PU239	QC	1.004e-2	1.02e-02	101.594 % Recovery	
11 CCV	0	0	OMSACT 241	QC	1.026e-2	1.12e-02	109.162 % Recovery	
12 CCB	0	0	OMSAC-QC PA231	QC	1	<3.30e-6		ug/mL
12 CCB	0	0	OMSAC-QC TH229	QC	1	<4.34e-6		ug/mL
12 CCB	0	0	OMSAC-QC TH230	QC	1	<3.21e-6		ug/mL
12 CCB	0	0	OMSAC-QC TH232	QC	1	<2.80e-6		ug/mL
12 CCB	0	0	OMSAC-QC NP237	QC	1	<1.48e-6		ug/mL
12 CCB	0	0	OMSAC-QC PU239	QC	1	<3.23e-6		ug/mL
12 CCB	0	0	OMSAC-QC PU240	QC	1	<2.18e-6		ug/mL
12 CCB	0	0	OMSAC-QC PU242	QC	1	<1.54e-6		ug/mL
12 CCB	0	0	OMSAC-QC 241	QC	1	<1.79e-6		ug/mL
12 CCB	0	0	OMSAC-QC 243	QC	1	<2.60e-6		ug/mL
12 CCB	0	0	OMSAC-QC 244	QC	1	<3.82e-6		ug/mL

Final page for worklist# 27741

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sam M. Pang 03/04/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27741

Analyst: KST Instrument: ICPMS1 Book# 11859

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSACT	QC		
2	ICB		@MSACTA1	SOLID		
3	BLNK-PREP		@MSACTA1	SOLID		
4	SAMPLE	S98T002847 0 A	@MSACTA1	SOLID	98000437	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
5	DUP	S98T002847 0 A	@MSACTA1	SOLID		
6	SAMPLE	S98T002926 0 A	@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
7	DUP	S98T002926 0 A	@MSACTA1	SOLID		
8	SAMPLE	S98T002927 0 A	@MSACTA1	SOLID	98000490	BY-105 (2)
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
9	DUP	S98T002927 0 A	@MSACTA1	SOLID		
10	SPK-POST	S98T002927 0 A	@MSACTA1	SOLID		
11	CCV		@MSACT	QC		
12	CCB		@MSACTA1	SOLID		

Final page for worklist # 27741

K3 Thomsdike 03-02-99
Analyst Signature Date

Analyst Signature Date

Data Entry Comments:

2847 .05-10-250-10 8241
2847-d " "
2926 .05-10 201
2926-d
2927
2927-d

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2927-S
cont of 8859 for spike
2904

03/02/99 17:30:02

Analysis Report

HNF-1663 REV. 0

Method: MS_TUNE

Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/02/99 17:29 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	205600.
Stddev	1778.
%RSD	.8646

#1	205800.
#2	203000.
#3	208300.
#4	205800.
#5	203900.
#6	207300.
#7	205800.
#8	207000.
#9	203100.
#10	206000.

File 030206.txt

S98T002847

2847-d

.05-10-.250-10
.05-10-.250-0
.05-108241 df
8241 df
201 df

S98T 002926

2926-d

2927

2927-d

2927-a

Uploaded & Validated by:

Saul M. Pang
03/04/99SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2905 TO 2935

Method: ACTINIDE Sample Name: cal blank

Operator: kjt

Comment: background

Run Time: 03/02/99 17:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2963	.3175	.3646	.3866	.3050
Stddev	.0485	.0365	.0328	.0332	.0381
%RSD	16.36	11.49	8.993	8.594	12.49

#1	.2920	.2953	.3543	.4101	.3084
#2	.3310	.3567	.3246	.4082	.2828
#3	.2174	.2998	.4151	.3821	.2734
#4	.3008	.3564	.3597	.4022	.2910
#5	.3402	.2792	.3691	.3306	.3691

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3098	.3652	.3095	.3489	.3180
Stddev	.0131	.0295	.0288	.0209	.0533
%RSD	4.243	8.080	9.289	5.992	16.76

#1	.3051	.3543	.3248	.3346	.2986
#2	.2893	.3632	.3246	.3246	.2732
#3	.3228	.4151	.2833	.3789	.2701
#4	.3172	.3564	.2747	.3564	.3597
#5	.3146	.3370	.3402	.3499	.3884

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3410	.3321
Stddev	.0240	.0358
%RSD	7.047	10.79

#1	.3314	.3543
#2	.3503	.2893
#3	.3360	.3294
#4	.3760	.3793
#5	.3113	.3081

HNF-1663 REV. 0

Sample Name: cal blank Run Time: 03/02/99 17:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	102.46
Stddev	1.24
%RSD	1.2102
#1	101.60
#2	103.72
#3	101.18
#4	101.95
#5	103.85

Analysis Report

HNF-1663 REV. 0

03/02/99 17:39:26

page 1

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.010 ppm cal

Run Time: 03/02/99 17:35 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2317	.3007	821.6	899.8	1.319
Stddev	.0221	.0263	16.8	15.4	.059
%RSD	9.520	8.729	2.049	1.716	4.449

#1	.2237	.3240	816.6	888.9	1.276
#2	.2113	.3308	830.5	913.5	1.259
#3	.2178	.2959	823.0	894.3	1.410
#4	.2395	.2835	796.4	883.7	1.318
#5	.2665	.2693	841.4	918.7	1.330

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.575	.5640	.2119	.2137	.2042
Stddev	.036	.0267	.0301	.0207	.0376
%RSD	2.291	4.728	14.23	9.690	18.40

#1	1.554	.5448	.1950	.2036	.2667
#2	1.540	.5504	.2446	.1890	.1723
#3	1.594	.5807	.2122	.2261	.1870
#4	1.629	.6028	.2367	.2422	.1844
#5	1.560	.5414	.1711	.2076	.2104

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2425	.2516
Stddev	.0444	.0341
%RSD	18.32	13.54

#1	.2753	.2782
#2	.1751	.2001
#3	.2680	.2596
#4	.2752	.2367
#5	.2188	.2833

Sample Name: cal std Run Time: 03/02/99 17:35

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	119.10
Stddev	1.80
%RSD	1.5155

#1	116.24
#2	119.91
#3	119.40
#4	121.11
#5	118.83

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.100 ppm

Run Time: 03/02/99 17:40 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3100	1.083	8068.	8650.	11.57
Stddev	.0266	.039	206.	215.	.26
%RSD	8.576	3.570	2.551	2.486	2.205
#1	.3045	1.107	7886.	8387.	11.28
#2	.2933	1.042	7933.	8469.	11.38
#3	.3188	1.079	7944.	8830.	11.63
#4	.3509	1.136	8345.	8870.	11.93
#5	.2823	1.053	8230.	8696.	11.63

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	14.01	2.383	.2534	.2421	.2569
Stddev	.26	.063	.0130	.0226	.0188
%RSD	1.877	2.637	5.143	9.336	7.328
#1	13.60	2.385	.2647	.2078	.2561
#2	14.13	2.417	.2534	.2306	.2876
#3	13.93	2.463	.2334	.2534	.2477
#4	14.29	2.350	.2502	.2617	.2560
#5	14.08	2.298	.2654	.2569	.2372

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2176	.2472
Stddev	.0192	.0150
%RSD	8.820	6.067
#1	.2334	.2647
#2	.2335	.2477
#3	.2249	.2249
#4	.2071	.2560
#5	.1892	.2428

Sample Name: cal std Run Time: 03/02/99 17:40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	117.05
Stddev	.77
%RSD	.65926
#1	117.12
#2	117.07
#3	117.09
#4	115.89
#5	118.06

Method: ACTINIDE Sample Name: icv

Operator: kjt

Comment: lcs std

Run Time: 03/02/99 17:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3020	.3431	835.7	14.05	1.548
Stddev	.0329	.0362	32.9	.31	.067
%RSD	10.89	10.56	3.938	2.200	4.328
#1	.2879	.3391	852.6	14.48	1.641
#2	.3366	.3860	791.7	13.87	1.594
#3	.2979	.2917	879.5	14.21	1.477
#4	.3310	.3310	832.7	14.00	1.522
#5	.2565	.3678	822.2	13.68	1.508

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	998.3	903.4	25.11	.4288	.3013
Stddev	29.8	15.5	.56	.0372	.0301
%RSD	2.984	1.720	2.231	8.682	10.00
#1	1045.	924.1	25.85	.3710	.2751
#2	979.3	891.1	25.53	.4231	.3150
#3	973.6	892.4	24.84	.4407	.2700
#4	983.4	893.3	24.77	.4734	.3434
#5	1010.	916.1	24.53	.4358	.3029

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	988.7	.2742
Stddev	14.4	.0207
%RSD	1.456	7.560
#1	1012.	.2975
#2	980.9	.2841
#3	991.2	.2421
#4	984.6	.2784
#5	974.7	.2689

Sample Name: icv Run Time: 03/02/99 17:55

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.03
Stddev	1.59
%RSD	1.4844
#1	104.21
#2	107.93
#3	107.41
#4	107.74
#5	107.85

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/02/99 18:17 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2776	.2714	.4307	.3649	.2881
Stddev	.0237	.0372	.0277	.0310	.0356
%RSD	8.519	13.71	6.426	8.506	12.35

#1	.2679	.2131	.3958	.3166	.2801
#2	.2611	.2551	.4471	.3571	.3361
#3	.2672	.2990	.4453	.3690	.2767
#4	.3193	.2922	.4067	.3976	.2410
#5	.2727	.2975	.4587	.3843	.3068

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2969	.3726	.2932	.3156	.3055
Stddev	.0183	.0228	.0215	.0256	.0232
%RSD	6.152	6.110	7.335	8.117	7.592

#1	.2679	.3562	.2588	.3197	.2892
#2	.2971	.3751	.3061	.2761	.3271
#3	.3117	.3976	.3085	.3340	.3117
#4	.3133	.3434	.3073	.3073	.2741
#5	.2944	.3905	.2851	.3409	.3254

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3121	.3114
Stddev	.0157	.0309
%RSD	5.025	9.937

#1	.3288	.3136
#2	.3271	.3001
#3	.3085	.3594
#4	.2922	.2741
#5	.3037	.3099

Sample Name: icb Run Time: 03/02/99 18:17

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	108.72
Stddev	2.58
%RSD	2.3732
#1	109.49
#2	111.07
#3	104.80
#4	110.66
#5	107.56

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: blnk-prep

Operator: kjt

Comment: 11 df

Run Time: 03/02/99 18:22 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3225	.3266	.5185	3.859	.3068
Stddev	.0206	.0529	.0374	.312	.0239
%RSD	6.390	16.20	7.221	8.097	7.805

#1	.3501	.3501	.5325	3.757	.3027
#2	.3329	.3884	.5734	4.339	.2738
#3	.3025	.3430	.4943	3.950	.3320
#4	.3021	.3021	.4761	3.500	.3283
#5	.3247	.2495	.5161	3.749	.2974

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2940	.4344	.3303	.2935	.3519
Stddev	.0153	.0499	.0328	.0508	.0525
%RSD	5.193	11.50	9.945	17.33	14.93

#1	.2845	.4559	.3720	.2480	.4231
#2	.2997	.4513	.3219	.3145	.3847
#3	.2729	.4979	.3098	.2729	.3246
#4	.3119	.3841	.2922	.2594	.2889
#5	.3008	.3828	.3555	.3725	.3384

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3414	.3437
Stddev	.0653	.0354
%RSD	19.12	10.30

#1	.3793	.3866
#2	.3551	.3699
#3	.3836	.3061
#4	.2265	.3447
#5	.3623	.3110

HNF-1663 REV. 0

Sample Name: blnk-prep Run Time: 03/02/99 18:22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	94.184
Stddev	5.101
%RSD	5.4156

#1	91.390
#2	90.104
#3	90.372
#4	101.53
#5	97.528

HNF-1663 REV 0

Method: ACTINIDE Sample Name: S98t002847

Operator: kjt

Comment: 8241 df

Run Time: 03/02/99 18:34 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2625	.2573	.5051	849.4	.2793
Stddev	.0314	.0272	.0516	23.6	.0228
%RSD	11.95	10.59	10.22	2.779	8.180

#1	.2972	.2669	.4237	814.7	.2779
#2	.2150	.2351	.5333	858.2	.2580
#3	.2587	.2816	.5489	863.1	.2558
#4	.2580	.2809	.4844	836.9	.3010
#5	.2836	.2222	.5350	873.9	.3040

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.604	.6067	.2652	.3109	.2810
Stddev	.037	.0321	.0284	.0319	.0129
%RSD	2.329	5.296	10.72	10.25	4.599

#1	1.590	.5833	.2394	.2999	.2779
#2	1.646	.6393	.3039	.3412	.2924
#3	1.561	.6380	.2385	.2989	.2759
#4	1.640	.5676	.2608	.2694	.2637
#5	1.581	.6051	.2836	.3449	.2952

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2770	.2884
Stddev	.0163	.0354
%RSD	5.886	12.29

#1	.2779	.3027
#2	.2523	.3154
#3	.2788	.2702
#4	.2780	.2350
#5	.2982	.3186

HNF-1663 REV. 0

Sample Name: s98t002847 Run Time: 03/02/99 18:34

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	116.74
Stddev	2.64
%RSD	2.2586

#1	121.15
#2	116.27
#3	115.98
#4	116.29
#5	114.03

Method: ACTINIDE Sample Name: s98t002847_d

Operator: kjt

Comment: 8241 df

Run Time: 03/02/99 18:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3064	.2892	.4999	1006.	.2889
Stddev	.0216	.0329	.0534	25.	.0238
%RSD	7.062	11.36	10.68	2.524	8.250

#1	.3355	.2494	.4929	991.2	.2791
#2	.2968	.3224	.4566	975.0	.2768
#3	.3183	.2891	.4497	1007.	.3096
#4	.3026	.3208	.5811	1042.	.3178
#5	.2786	.2641	.5194	1015.	.2612

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.814	.6373	.3007	.2957	.3149
Stddev	.035	.0388	.0256	.0284	.0385
%RSD	1.949	6.085	8.501	9.594	12.23

#1	1.802	.5939	.3029	.2999	.2554
#2	1.789	.5992	.2825	.2597	.3025
#3	1.819	.6746	.3066	.3008	.3446
#4	1.873	.6718	.3389	.2814	.3208
#5	1.788	.6471	.2728	.3366	.3511

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2915	.3114
Stddev	.0238	.0299
%RSD	8.168	9.599

#1	.2761	.3266
#2	.3110	.2939
#3	.2570	.2891
#4	.3057	.3571
#5	.3076	.2902

Sample Name: s98t002847_d Run Time: 03/02/99 18:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	113.65
Stddev	2.55
%RSD	2.2427

#1	112.26
#2	116.82
#3	114.14
#4	110.14
#5	114.87

Method: ACTINIDE Sample Name: s98t002926

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 18:45 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2363	.2522	20.62	295.9	.2890
Stddev	.0344	.0125	.36	3.5	.0168
%RSD	14.58	4.952	1.741	1.183	5.813

#1	.2514	.2599	20.58	292.5	.2768
#2	.2010	.2615	20.02	291.7	.2946
#3	.2844	.2450	20.75	298.9	.3126
#4	.2056	.2611	20.96	298.8	.2917
#5	.2391	.2336	20.79	297.4	.2693

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.7157	.4741	.2661	.2872	.2760
Stddev	.0300	.0486	.0198	.0240	.0259
%RSD	4.198	10.25	7.447	8.338	9.374

#1	.6723	.4011	.2797	.2994	.2542
#2	.7240	.4597	.2918	.3083	.3166
#3	.7181	.5012	.2619	.2844	.2760
#4	.7083	.4778	.2417	.2472	.2805
#5	.7557	.5304	.2556	.2968	.2528

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2853	.2644
Stddev	.0329	.0329
%RSD	11.52	12.45

#1	.3390	.2825
#2	.2588	.2615
#3	.2872	.2563
#4	.2833	.2167
#5	.2583	.3050

Sample Name: s98t002926 Run Time: 03/02/99 18:45

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	119.75
Stddev	1.52
%RSD	1.2677

#1	117.99
#2	121.09
#3	118.37
#4	120.00
#5	121.30

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: s98t002926_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 18:49 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2342	.2515	5.431	238.0	.2498
Stddev	.0431	.0246	.102	2.8	.0113
%RSD	18.41	9.790	1.872	1.168	4.510

#1	.2069	.2311	5.458	242.5	.2472
#2	.2374	.2617	5.507	238.1	.2428
#3	.2148	.2390	5.494	237.8	.2363
#4	.3077	.2355	5.255	235.7	.2622
#5	.2041	.2901	5.439	235.7	.2606

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.6201	.3848	.2557	.2370	.2762
Stddev	.0209	.0416	.0205	.0167	.0133
%RSD	3.363	10.82	8.032	7.026	4.807

#1	.6180	.3816	.2822	.2580	.2607
#2	.6368	.4155	.2266	.2482	.2887
#3	.6069	.4189	.2632	.2309	.2632
#4	.5941	.3158	.2596	.2328	.2863
#5	.6447	.3922	.2471	.2149	.2820

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2676	.2607
Stddev	.0266	.0380
%RSD	9.959	14.59

#1	.2768	.2768
#2	.2590	.3049
#3	.2282	.2282
#4	.2729	.2141
#5	.3008	.2794

HNF-1663 REV. 0

Sample Name: s98t002926_d Run Time: 03/02/99 18:49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	124.08
Stddev	.37
%RSD	.29474

#1	124.05
#2	123.54
#3	124.13
#4	124.57
#5	124.09

Analysis Report

HNF-1663 REV. 0

03/02/99 18:56:38

page 1

Method: ACTINIDE Sample Name: s98t002927

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 18:53 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2588	.2722	1.588	207.0	.2558
Stddev	.0160	.0152	.075	4.2	.0276
%RSD	6.163	5.571	4.695	2.049	10.81

#1	.2608	.2951	1.657	203.3	.2687
#2	.2451	.2639	1.565	206.6	.2370
#3	.2410	.2572	1.587	206.8	.2626
#4	.2677	.2651	1.476	204.2	.2200
#5	.2795	.2795	1.655	214.1	.2906

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.6008	.3635	.2683	.2557	.2604
Stddev	.0223	.0494	.0393	.0432	.0171
%RSD	3.713	13.58	14.64	16.90	6.573

#1	.6165	.3478	.2292	.2371	.2608
#2	.6302	.3016	.3205	.2989	.2397
#3	.5957	.3493	.2708	.1922	.2870
#4	.5752	.3844	.2306	.2598	.2598
#5	.5867	.4345	.2906	.2906	.2546

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2623	.2448
Stddev	.0246	.0116
%RSD	9.382	4.748

#1	.3056	.2582
#2	.2585	.2397
#3	.2491	.2518
#4	.2465	.2280
#5	.2518	.2463

Sample Name: s98t002927 Run Time: 03/02/99 18:53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	123.92
Stddev	2.39
%RSD	1.9278
#1	126.52
#2	123.77
#3	123.11
#4	125.75
#5	120.45

Method: ACTINIDE

Sample Name: s98t002927_d

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 18:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2698	.2443	1.585	223.4	.2606
Stddev	.0492	.0325	.071	6.4	.0217
%RSD	18.23	13.32	4.502	2.847	8.336

#1	.2952	.2817	1.590	226.5	.2383
#2	.2496	.2523	1.519	227.7	.2686
#3	.1966	.2047	1.508	212.2	.2370
#4	.3256	.2171	1.636	224.9	.2849
#5	.2822	.2660	1.672	225.9	.2741

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.6018	.4003	.2876	.2822	.2600
Stddev	.0469	.0425	.0362	.0195	.0268
%RSD	7.787	10.62	12.59	6.912	10.31

#1	.5714	.4144	.2329	.3114	.2952
#2	.6809	.3364	.3120	.2821	.2306
#3	.6087	.4148	.2936	.2694	.2532
#4	.5752	.4504	.3256	.2605	.2415
#5	.5726	.3854	.2741	.2877	.2795

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2497	.2627
Stddev	.0326	.0264
%RSD	13.07	10.05

#1	.2546	.2248
#2	.1953	.2604
#3	.2694	.2532
#4	.2496	.2903
#5	.2795	.2850

HNF-1663 REV. 0

Sample Name: s98t002927_d Run Time: 03/02/99 18:57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	123.08
Stddev	.39
%RSD	.31611

#1	123.08
#2	122.88
#3	123.75
#4	122.85
#5	122.83

Analysis Report

HNF-1663 REV.0

03/02/99 19:04:32

page 1

Method: ACTINIDE

Sample Name: s98t002927_s

should be - a

Operator: kjt

Comment: 201 df

Run Time: 03/02/99 19:00 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2643	.4164	1628.	212.2	2.389
Stddev	.0175	.0155	32.	1.6	.060
%RSD	6.603	3.724	1.953	.7699	2.511

#1	.2645	.4200	1584.	211.3	2.386
#2	.2604	.4368	1667.	214.6	2.439
#3	.2667	.4139	1637.	210.1	2.428
#4	.2894	.4177	1611.	212.5	2.288
#5	.2405	.3935	1640.	212.4	2.402

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1996.	1743.	47.03	.4805	.2518
Stddev	17.	35.	1.71	.0274	.0129
%RSD	.8697	1.982	3.628	5.702	5.130

#1	1967.	1704.	44.09	.4718	.2536
#2	2009.	1749.	48.31	.4396	.2688
#3	1994.	1797.	47.74	.4861	.2583
#4	2006.	1730.	47.02	.4914	.2375
#5	2005.	1733.	47.96	.5137	.2405

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1946.	.2744
Stddev	37.	.0279
%RSD	1.913	10.18

#1	1904.	.2809
#2	1992.	.2912
#3	1974.	.3083
#4	1917.	.2430
#5	1941.	.2487

Analysis Report

should be a 03/02/99 19:04:32

page 2

Sample Name: s98t002927_s Run Time: 03/02/99 19:00

HNF-1663 REV. 0

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	121.07
Stddev	1.45
%RSD	1.2012
#1	122.23
#2	119.04
#3	120.00
#4	122.09
#5	121.98

Analysis Report

03/02/99 19:13:31

page 1

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccv

Operator: kjt

Comment: 1sc std

Run Time: 03/02/99 19:09 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2793	.3643	862.7	14.69	1.456
Stddev	.0217	.0489	17.1	.15	.081
%RSD	7.762	13.42	1.988	1.053	5.584

#1	.2644	.3526	862.2	14.81	1.425
#2	.3070	.3391	861.9	14.79	1.488
#3	.2620	.4347	835.9	14.72	1.566
#4	.2986	.3071	882.5	14.42	1.459
#5	.2644	.3880	870.9	14.69	1.345

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1045.	935.0	25.62	.4136	.2931
Stddev	9.	12.0	.47	.0126	.0201
%RSD	.8542	1.283	1.848	3.051	6.848

#1	1048.	947.4	25.64	.4009	.2957
#2	1040.	923.5	25.65	.4180	.3099
#3	1033.	921.8	25.62	.4261	.2965
#4	1051.	937.0	24.92	.4237	.2588
#5	1056.	945.5	26.26	.3995	.3046

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1052.	.3200
Stddev	19.	.0427
%RSD	1.768	13.35

#1	1050.	.3043
#2	1050.	.3333
#3	1068.	.2533
#4	1023.	.3498
#5	1069.	.3592

Sample Name: ccv   Run Time: 03/02/99 19:09

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	116.05
Stddev	1.32
%RSD	1.1342

#1	117.23
#2	114.02
#3	115.79
#4	117.22
#5	115.99

Analysis Report

HNF-1663 REV. 0

03/02/99 19:29:50

page 1

Method: ACTINIDE Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/02/99 19:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2699	.2955	.4080	.3945	.2990
Stddev	.0111	.0382	.0264	.0330	.0338
%RSD	4.097	12.91	6.462	8.377	11.31

#1	.2665	.2752	.3910	.3505	.3273
#2	.2567	.3216	.3752	.4175	.2623
#3	.2820	.3473	.4066	.4215	.3384
#4	.2808	.2528	.4269	.4156	.2977
#5	.2637	.2806	.4405	.3676	.2694

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2984	.4115	.2883	.3032	.2856
Stddev	.0596	.0400	.0340	.0199	.0391
%RSD	19.99	9.713	11.79	6.577	13.68

#1	.3534	.4316	.2491	.3215	.2838
#2	.2370	.3865	.2595	.2708	.3018
#3	.2701	.4393	.3295	.3176	.3146
#4	.3707	.4465	.3117	.3033	.3089
#5	.2609	.3535	.2918	.3030	.2189

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2863	.3101
Stddev	.0375	.0147
%RSD	13.09	4.730

#1	.2462	.3070
#2	.3103	.3131
#3	.2938	.3324
#4	.3314	.2921
#5	.2497	.3058

Analysis Report

HNF-1663 REV.0

Sample Name: ccb   Run Time: 03/02/99 19:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	116.61
Stddev	2.85
%RSD	2.4413
#1	115.09
#2	118.16
#3	112.30
#4	118.69
#5	118.80

LABCORE Completed Worklist Report for Worklist# 27742

Analyst: kjt

Instrument: ICPMS1

Book#: 48B57

Method: LA-506-101 Rev/Mod A-0

49B57

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSPR-QC TC99	QC	0.0205	1.87e-02	91.220 % Recovery	
1	ICV	0	QMSPR-QC SN117	QC	0.01925	2.15e-02	111.688 % Recovery	
1	ICV	0	QMSPR-QC CS133	QC	0.020	1.95e-02	97.500 % Recovery	
2	ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
2	ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
2	ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
2	ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL
3	BLNK-PREP	0	QMSPR-A1 TC99-A1	SOLID	1	3.66e-04	3.660e-004	ug/g
3	BLNK-PREP	0	QMSPR-A1 SN117-A1	SOLID	1	4.60e-03	0.005	ug/g
3	BLNK-PREP	0	QMSPR-A1 SN126-A1	SOLID	1	<1.65e-3		ug/g
3	BLNK-PREP	0	QMSPR-A1 CS133-A1	SOLID	1	2.42e-05	2.420e-005	ug/g
3	BLNK-PREP	0	QMSPR-A1 CS135-A1	SOLID	1	4.62e-05	4.620e-005	ug/g
3	BLNK-PREP	0	QMSPR-A1 CS137-A1	SOLID	1	7.92e-05	7.920e-005	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 TC99-A1	SOLID	N/A	< 8.431e+00	8.431	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 SN117-A1	SOLID	N/A	< 6.323e+02	632.400	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 SN126-A1	SOLID	N/A	< 6.323e+02	632.400	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 CS133-A1	SOLID	N/A	< 8.431e+00	8.431	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 CS135-A1	SOLID	N/A	1.012e+01	8.431	ug/g
4	SAMPLE	S98T002847	0 A QMSPR-A1 CS137-A1	SOLID	N/A	2.361e+01	8.431	ug/g
5	DUP	S98T002847	0 A QMSPR-A1 TC99-A1	SOLID	<8.43e0	<7.66e0		RPD
5	DUP	S98T002847	0 A QMSPR-A1 SN117-A1	SOLID	<6.32e2	<5.75e2		RPD
5	DUP	S98T002847	0 A QMSPR-A1 SN126-A1	SOLID	<6.32e2	<5.75e2		RPD
5	DUP	S98T002847	0 A QMSPR-A1 CS133-A1	SOLID	<8.43e0	<7.66e0		RPD
5	DUP	S98T002847	0 A QMSPR-A1 CS135-A1	SOLID	1.01e+01	1.34e+01	28.085	RPD
5	DUP	S98T002847	0 A QMSPR-A1 CS137-A1	SOLID	2.36e+01	2.87e+01	19.503	RPD
6	SAMPLE	S98T002926	0 A QMSPR-A1 TC99-A1	SOLID	N/A	8.438e-01	7.96e-002	ug/g
6	SAMPLE	S98T002926	0 A QMSPR-A1 SN117-A1	SOLID	N/A	< 5.970e+00	5.970	ug/g
6	SAMPLE	S98T002926	0 A QMSPR-A1 SN126-A1	SOLID	N/A	< 5.970e+00	5.970	ug/g
6	SAMPLE	S98T002926	0 A QMSPR-A1 CS133-A1	SOLID	N/A	1.079e+00	7.96e-002	ug/g
6	SAMPLE	S98T002926	0 A QMSPR-A1 CS135-A1	SOLID	N/A	4.577e-01	7.96e-002	ug/g
6	SAMPLE	S98T002926	0 A QMSPR-A1 CS137-A1	SOLID	N/A	1.011e+00	7.96e-002	ug/g
7	DUP	S98T002926	0 A QMSPR-A1 TC99-A1	SOLID	8.44e-01	7.33e-01	14.077	RPD
7	DUP	S98T002926	0 A QMSPR-A1 SN117-A1	SOLID	<5.97e0	<6.04e0		RPD
7	DUP	S98T002926	0 A QMSPR-A1 SN126-A1	SOLID	<5.97e0	<6.04e0		RPD
7	DUP	S98T002926	0 A QMSPR-A1 CS133-A1	SOLID	1.08e+00	9.18e-01	16.216	RPD
7	DUP	S98T002926	0 A QMSPR-A1 CS135-A1	SOLID	4.58e-01	3.67e-01	22.061	RPD
7	DUP	S98T002926	0 A QMSPR-A1 CS137-A1	SOLID	1.01e+00	8.42e-01	18.143	RPD
8	SAMPLE	S98T002927	0 A QMSPR-A1 TC99-A1	SOLID	N/A	3.544e-01	7.88e-002	ug/g
8	SAMPLE	S98T002927	0 A QMSPR-A1 SN117-A1	SOLID	N/A	< 5.907e+00	5.907	ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27742

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
8 SAMPLE	S98T002927	0 A	QMSPR-A1 SN126-A1 SOLID		N/A <	5.907e+00	5.907	ug/g
8 SAMPLE	S98T002927	0 A	QMSPR-A1 CS133-A1 SOLID		N/A	4.686e-01	7.88e-002	ug/g
8 SAMPLE	S98T002927	0 A	QMSPR-A1 CS135-A1 SOLID		N/A	3.032e-01	7.88e-002	ug/g
8 SAMPLE	S98T002927	0 A	QMSPR-A1 CS137-A1 SOLID		N/A	7.010e-01	7.88e-002	ug/g
9 DUP	S98T002927	0 A	QMSPR-A1 TC99-A1 SOLID		3.54e-01	3.68e-01	3.878	RPD
9 DUP	S98T002927	0 A	QMSPR-A1 SN117-A1 SOLID		<5.91e0	<5.75e0		RPD
9 DUP	S98T002927	0 A	QMSPR-A1 SN126-A1 SOLID		<5.91e0	<5.75e0		RPD
9 DUP	S98T002927	0 A	QMSPR-A1 CS133-A1 SOLID		4.69e-01	4.37e-01	7.064	RPD
9 DUP	S98T002927	0 A	QMSPR-A1 CS135-A1 SOLID		3.03e-01	2.11e-01	35.798	RPD
9 DUP	S98T002927	0 A	QMSPR-A1 CS137-A1 SOLID		7.01e-01	5.32e-01	27.413	RPD
10 SPK-POST	S98T002927	0 A	QMSPR-A1 TC99-A1 SOLID		0.0219	2.00e-02	91.324	% Recovery
10 SPK-POST	S98T002927	0 A	QMSPR-A1 SN117-A1 SOLID		0.0193	1.90e-02	98.446	% Recovery
10 SPK-POST	S98T002927	0 A	QMSPR-A1 CS133-A1 SOLID		0.0201	2.02e-02	100.498	% Recovery
11 CCV	0	0	QMSPR-QC TC99	QC	0.0205	1.88e-02	91.707	% Recovery
11 CCV	0	0	QMSPR-QC SN117	QC	0.01925	2.13e-02	110.649	% Recovery
11 CCV	0	0	QMSPR-QC CS133	QC	0.0200	1.97e-02	98.500	% Recovery
12 CCB	0	0	QMSPR-QC TC99	QC	1	<2.00e-6		ug/mL
12 CCB	0	0	QMSPR-QC SN117	QC	1	<1.50e-4		ug/mL
12 CCB	0	0	QMSPR-QC SN126	QC	1	<1.50e-4		ug/mL
12 CCB	0	0	QMSPR-QC CS133	QC	1	<2.00e-6		ug/mL
12 CCB	0	0	QMSPR-QC CS135	QC	1	<2.00e-6		ug/mL
12 CCB	0	0	QMSPR-QC CS137	QC	1	<2.00e-6		ug/mL

Final page for worklist# 27742

Analyst Signature

Date

Analyst Signature

Date

Paul B. Parag 03/08/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 27742

Analyst: KST Instrument: ICPMS1 Book# 48857Method: LA-506-101 Rev/Mod A-O 49857

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC QC			
2 ICB			@MSPR-QC QC			
3 BLNK-PREP			@MSPR-A1 SOLID			
4 SAMPLE	S98T002847 0 A		@MSPR-A1 SOLID		98000437 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
5 DUP	S98T002847 0 A		@MSPR-A1 SOLID			
6 SAMPLE	S98T002926 0 A		@MSPR-A1 SOLID		98000490 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
7 DUP	S98T002926 0 A		@MSPR-A1 SOLID			
8 SAMPLE	S98T002927 0 A		@MSPR-A1 SOLID		98000490 BY-105 (2)	
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, I129-A1, SN117-A1, SN126-A1, TC99-A1					
9 DUP	S98T002927 0 A		@MSPR-A1 SOLID			
10 SPK-POST	S98T002927 0 A		@MSPR-A1 SOLID			
11 CCV			@MSPR-QC QC			
12 CCB			@MSPR-QC QC			

Final page for worklist # 27742

K. S. Thendike 03-07-99
Analyst Signature Date_____
Analyst Signature Date

Data Entry Comments:

Wrong df for S98T002847 on Scan in
Computer should be 20301Uploaded & Validated by:
Saul M. Pang
03/08/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

HNF-1663 REV. 0

Method: MS_TUNE

Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 03/07/99 09:13 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	176900.
Stddev	1603.
%RSD	.9066

#1	177800.
#2	177600.
#3	175800.
#4	176900.
#5	175500.
#6	175900.
#7	174800.
#8	180500.
#9	176800.
#10	177100.

cf
3/7/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2939 TO 2961.

file 0307a.txt

Scan pr 27742

pr 2847

pr 2926

pr 2927

Blk prep

1-10

11 df

20301 03-07-99

S98T002847

10-1-10

40801 df

-d

S98T002926

05-10

201 df

2926-d

2927

2927-d

-a

-a

Spike

1 ml of msprv spk 44357

25 ml of 10 ppm Sn 990119-1

Analysis Report

HNF-1663 REV. 0

03/07/99 09:17:37

page 1

Method: PRV

Sample Name: Blank

Operator:

Comment:

Run Time: 03/07/99 09:15 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.055	49.30	1.295	16.66	1.522
Stddev	.220	4.28	.183	.42	.100
%RSD	10.68	8.683	14.14	2.538	6.544

#1	2.060	45.00	1.016	17.41	1.383
#2	1.929	48.50	1.454	16.52	1.566
#3	2.058	48.50	1.458	16.38	1.629
#4	1.822	48.00	1.318	16.51	1.458
#5	2.405	56.50	1.231	16.49	1.575

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.320	2.477	3.112	12.59
Stddev	.116	.098	.248	.55
%RSD	4.987	3.967	7.973	4.337

#1	2.258	2.512	3.048	12.31
#2	2.377	2.572	3.020	13.00
#3	2.487	2.344	2.801	13.15
#4	2.187	2.551	3.224	11.80
#5	2.291	2.405	3.465	12.69

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.678
Stddev	.194
%RSD	1.0954

#1	17.716
#2	17.883
#3	17.492
#4	17.836
#5	17.461

Method: PRV

Sample Name: PRV CAL

Operator:

Comment:

Run Time: 03/07/99

09:18 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Cs[133]
Line	99/pulse	133/pulse
Units	Cts/S	Cts/S
Avg	22620.	19570.
Stddev	217.	232.
%RSD	.9576	1.188

#1	22440.	19610.
#2	22890.	19840.
#3	22770.	19260.
#4	22610.	19710.
#5	22370.	19410.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.573
Stddev	.156
%RSD	.88725

#1	17.388
#2	17.522
#3	17.697
#4	17.490
#5	17.768

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV1 CAL

Operator:

Comment:

Run Time: 03/07/99 09:23 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Ru[102]	Te[125]	Cs[135]	Cs[137]
Line	102/pulse	125/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	627300.	1516.	6545.	11460.
Stddev	2943.	52.	218.	358.
%RSD	.4692	3.426	3.333	3.124

#1	628000.	1443.	6266.	10850.
#2	628000.	1488.	6494.	11510.
#3	624600.	1573.	6488.	11530.
#4	624500.	1524.	6612.	11730.
#5	631600.	1554.	6865.	11690.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	70530.
Stddev	2399.
%RSD	3.401

#1	66630.
#2	69780.
#3	72040.
#4	72170.
#5	72030.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.774
Stddev	.548
%RSD	3.0845

#1	18.687
#2	17.892
#3	17.416
#4	17.469
#5	17.408

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1147.9465	2.0550	1.0000000	03/07/99 09:19:51
Ru[102]	6272.8159	49.3000	1.0000000	03/07/99 09:24:41
Te[125]	15.1512	1.2953	1.0000000	03/07/99 09:24:41
Sn[117]	952.5544	16.6644	1.0000000	03/07/99 09:17:35
Sn[126]	953.3411	1.5221	1.0000000	03/07/99 09:17:35
Cs[133]	978.2415	2.3197	1.0000000	03/07/99 09:19:51
Cs[135]	991.2658	2.4769	1.0000000	03/07/99 09:24:41
Cs[137]	1023.1808	3.1115	1.0000000	03/07/99 09:24:41
Ba[138]	705.1716	12.5885	1.0000000	03/07/99 09:24:41

Method: PRV

Sample Name: PRV-SN CAL

Operator:

Comment:

Run Time: 03/07/99 09:30 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Sn[117]	Sn[126]
Line	117/pulse	126/pulse
Units	Cts/S	Cts/S
Avg	18250.	18250.
Stdddev	364.	364.
%RSD	1.995	1.995

#1	18330.	18330.
#2	18270.	18270.
#3	17850.	17850.
#4	18790.	18790.
#5	17990.	17990.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.978
Stdddev	.333
%RSD	1.9615

#1	17.207
#2	16.943
#3	17.088
#4	16.419
#5	17.233

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1147.9465	2.0550	1.0000000	03/07/99 09:19:51
Ru[102]	6272.8159	49.3000	1.0000000	03/07/99 09:24:41
Te[125]	15.1512	1.2953	1.0000000	03/07/99 09:24:41
Sn[117]	947.0891	16.6644	1.0000000	03/07/99 09:30:56
Sn[126]	947.8757	1.5221	1.0000000	03/07/99 09:30:56
Cs[133]	978.2415	2.3197	1.0000000	03/07/99 09:19:51
Cs[135]	991.2658	2.4769	1.0000000	03/07/99 09:24:41
Cs[137]	1023.1808	3.1115	1.0000000	03/07/99 09:24:41
Ba[138]	705.1716	12.5885	1.0000000	03/07/99 09:24:41

Analysis Report

Method: PRV

Sample Name: blank

Operator: kjt

Comment:

Run Time: 03/07/99 09:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0013	.0034	-.0006	.0009	.0004
Stddev	.0005	.0011	.0097	.0004	.0002
%RSD	34.23	33.65	1658.	44.41	52.74

#1	.0019	.0019	.0050	.0014	.0004
#2	.0016	.0041	-.0172	.0004	.0004
#3	.0015	.0044	-.0005	.0009	.0007
#4	.0009	.0040	.0030	.0008	.0007
#5	.0008	.0023	.0068	.0013	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0002	.0000	.0009
Stddev	.0003	.0003	.000	.0011
%RSD	42.22	148.4	583.6	119.8

#1	.0010	-.0004	.0002	.0016
#2	.0008	-.0004	.0001	-.0001
#3	.0004	-.0002	-.0002	.0000
#4	.0005	-.0002	.0001	.0006
#5	.0004	.0002	-.0004	.0024

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.727
Stddev	.214
%RSD	1.2051

#1	17.505
#2	17.871
#3	17.481
#4	17.890
#5	17.887

Analysis Report

HNF-1663 REV. 0

03/07/99 09:45:15

page 1

Method: PRV

Sample Name: icv

Operator: kjt

Comment: msprv lcs

Run Time: 03/07/99 09:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.73	.0017	-.0023	1.621	.0004
Stddev	.20	.0014	.0064	.064	.0001
%RSD	1.047	80.46	276.9	3.973	28.58

#1	19.04	.0023	-.0070	1.570	.0005
#2	18.57	.0035	.0065	1.562	.0005
#3	18.61	.0021	.0011	1.593	.0005
#4	18.81	.0000	-.0027	1.694	.0003
#5	18.63	.0007	-.0096	1.686	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.47	.0073	.0120	.1131
Stddev	.29	.0007	.0004	.0037
%RSD	1.469	9.889	3.630	3.257

#1	19.91	.0084	.0123	.1169
#2	19.38	.0071	.0120	.1101
#3	19.53	.0076	.0119	.1085
#4	19.13	.0070	.0126	.1136
#5	19.41	.0065	.0114	.1162

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.851
Stddev	.118
%RSD	.66009

#1	17.652
#2	17.931
#3	17.904
#4	17.931
#5	17.837

Method: PRV Sample Name: icv

Operator: kjt

Comment: 19.25 ppb sn

Run Time: 03/07/99 09:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.3420	.0007	-.0005	21.48	.0005
Stddev	.4482	.0005	.0191	.24	.0002
%RSD	131.1	73.27	4074.	1.113	32.71
#1	1.127	.0010	-.0077	21.30	.0002
#2	.2970	.0004	.0249	21.77	.0005
#3	.1354	.0012	.0122	21.65	.0004
#4	.0853	.0000	-.0082	21.50	.0006
#5	.0652	.0008	-.0236	21.19	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0138	.0296	.0475	.4431
Stddev	.0010	.0012	.0017	.0135
%RSD	7.489	4.108	3.496	3.036
#1	.0150	.0302	.0453	.4235
#2	.0139	.0294	.0469	.4496
#3	.0141	.0296	.0498	.4599
#4	.0140	.0311	.0474	.4436
#5	.0122	.0278	.0481	.4390

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.857
Stddev	.218
%RSD	1.3741

#1	16.111
#2	15.841
#3	15.535
#4	15.803
#5	15.995

Analysis Report

HNF-1663 REV. 0

03/07/99 10:14:11

page 1

Method: PRV

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 03/07/99 10:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0018	-.0003	.0027	.0002
Stddev	.0004	.0012	.0090	.0002	.0001
%RSD	132.6	68.98	3089.	6.678	64.85
#1	.0010	.0028	-.0068	.0027	.0001
#2	.0000	.0000	-.0078	.0026	.0002
#3	.0001	.0027	.0019	.0028	.0001
#4	.0004	.0010	.0142	.0031	.0004
#5	.0001	.0022	-.0029	.0027	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0012	-.0001	-.0001	.0003
Stddev	.0003	.0003	.0002	.0009
%RSD	30.04	210.2	203.3	261.9
#1	.0017	-.0005	.0000	-.0000
#2	.0012	-.0001	-.0002	.0001
#3	.0010	.0000	-.0003	-.0006
#4	.0007	-.0003	-.0002	.0005
#5	.0012	.0002	.0002	.0018

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	17.973
Stddev	.116
%RSD	.64633

#1	18.040
#2	17.844
#3	18.122
#4	17.870
#5	17.990

Method: PRV

Sample Name: blnk-prep

Operator: kjt

Comment: 11 df

Run Time: 03/07/99 10:14 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0333	.0012	.0011	.4185	.0004
Stddev	.0123	.0009	.0136	.0129	.0002
%RSD	36.91	75.64	1210.	3.077	56.31
#1	.0511	.0011	.0125	.4088	.0003
#2	.0393	.0027	.0147	.4217	.0003
#3	.0309	.0011	.0000	.4379	.0002
#4	.0254	.0009	-.0192	.4192	.0007
#5	.0198	.0003	-.0024	.4049	.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0022	.0042	.0072	.0689
Stddev	.0004	.0007	.0009	.0026
%RSD	18.52	17.66	11.94	3.752
#1	.0029	.0043	.0064	.0688
#2	.0021	.0038	.0078	.0681
#3	.0022	.0032	.0078	.0716
#4	.0017	.0050	.0080	.0709
#5	.0023	.0047	.0062	.0650

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	16.292
Stddev	.300
%RSD	1.8413

#1	16.503
#2	16.144
#3	16.208
#4	15.923
#5	16.680

Method: PRV₂₀₃₀₁

Sample Name: s98t002847

Operator: kjt

Comment: 40581 df

Run Time: 03/07/99 10:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0002	.0033	.0028	.0022	.0004
Stddev	.0002	.0006	.0133	.0015	.0002
%RSD	141.1	19.35	474.9	66.54	40.45
#1	.0002	.0026	-.0088	.0033	.0006
#2	.0003	.0026	.0256	.0038	.0003
#3	.0004	.0039	.0002	.0019	.0002
#4	-.0002	.0036	-.0040	.0020	.0005
#5	.0002	.0035	.0010	.0000	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0015	.0024	.0056	.0383
Stddev	.0003	.0005	.0004	.0013
%RSD	18.44	20.38	8.093	3.431
#1	.0016	.0028	.0054	.0390
#2	.0011	.0019	.0062	.0377
#3	.0018	.0019	.0058	.0366
#4	.0014	.0029	.0050	.0401
#5	.0015	.0024	.0054	.0382

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.871
Stddev	.132
%RSD	.73685

#1	18.063
#2	17.830
#3	17.712
#4	17.823
#5	17.928

Analysis Report 303/8/91

HNF-1663 REV. 0

03/07/99 10:26:44

page 1

Method: PRV₂₀₃₀₁

Sample Name: s98t002847_d

Operator: kjt

Comment: 40501 df

Run Time: 03/07/99 10:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0039	.0056	.0008	.0003
Stddev	.0003	.0008	.0167	.0008	.0002
%RSD	197.9	19.87	297.2	105.9	59.81

#1	.0002	.0033	.0165	.0015	.0002
#2	.0004	.0043	.0186	.0016	.0001
#3	.0000	.0028	-.0122	.0004	.0004
#4	.0002	.0042	.0183	.0006	.0002
#5	-.0003	.0047	-.0132	-.0003	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0016	.0035	.0075	.0557
Stddev	.0004	.0003	.0004	.0015
%RSD	22.94	8.487	5.165	2.620

#1	.0011	.0035	.0076	.0575
#2	.0016	.0039	.0078	.0546
#3	.0020	.0035	.0071	.0555
#4	.0018	.0031	.0079	.0569
#5	.0013	.0036	.0071	.0540

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.046
Stddev	.171
%RSD	.94810

#1	17.791
#2	18.067
#3	17.999
#4	18.116
#5	18.257

Method: PRV   Sample Name: s98t002926   Operator: kjt
 Comment: 201 df
 Run Time: 03/07/99 10:27 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0212	.0213	-.0002	.0271	.0005
Stddev	.0009	.0017	.0061	.0012	.0003
%RSD	4.286	7.932	2563.	4.428	50.28

#1	.0212	.0234	.0001	.0268	.0007
#2	.0214	.0220	-.0085	.0263	.0004
#3	.0210	.0218	-.0003	.0282	.0004
#4	.0199	.0192	-.0011	.0258	.0009
#5	.0224	.0199	.0086	.0286	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0271	.0115	.0254	.0433
Stddev	.0009	.0004	.0012	.0013
%RSD	3.433	3.414	4.611	3.098

#1	.0285	.0113	.0244	.0445
#2	.0278	.0119	.0273	.0436
#3	.0268	.0118	.0250	.0430
#4	.0263	.0109	.0254	.0411
#5	.0264	.0115	.0246	.0443

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.188
Stddev	.142
%RSD	.78211

#1	18.122
#2	17.996
#3	18.204
#4	18.380
#5	18.239

Method: PRV

Sample Name: s98t002926_d

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 10:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0182	.0197	-.0095	.0388	.0005
Stddev	.0009	.0011	.0130	.0011	.0003
%RSD	4.751	5.674	137.2	2.861	63.25
#1	.0196	.0190	-.0236	.0390	.0002
#2	.0182	.0210	.0112	.0378	.0009
#3	.0173	.0193	-.0122	.0406	.0009
#4	.0178	.0208	-.0154	.0379	.0003
#5	.0180	.0184	-.0073	.0388	.0003

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0228	.0091	.0209	.0301
Stddev	.0010	.0003	.0008	.0015
%RSD	4.429	3.261	3.798	5.025
#1	.0233	.0089	.0210	.0318
#2	.0225	.0087	.0211	.0315
#3	.0241	.0092	.0218	.0289
#4	.0227	.0093	.0210	.0300
#5	.0214	.0094	.0196	.0284

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.746
Stddev	.197
%RSD	1.0517

#1	18.664
#2	18.769
#3	18.468
#4	18.830
#5	18.999

Method: PRV

Sample Name: s98t002927

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 10:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0090	.0098	-.0125	.0368	.0002
Stddev	.0007	.0016	.0101	.0013	.0002
%RSD	7.334	15.97	80.77	3.479	89.42

#1	.0099	.0088	-.0066	.0366	.0002
#2	.0091	.0077	-.0249	.0353	.0004
#3	.0092	.0101	-.0191	.0369	.0002
#4	.0082	.0106	.0008	.0388	.0002
#5	.0085	.0117	-.0128	.0364	-.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0119	.0077	.0178	.0617
Stddev	.0006	.0004	.0006	.0018
%RSD	4.961	4.825	3.553	2.857

#1	.0121	.0073	.0184	.0605
#2	.0121	.0076	.0180	.0610
#3	.0113	.0076	.0184	.0608
#4	.0126	.0078	.0174	.0614
#5	.0112	.0083	.0170	.0648

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	18.459
Stddev	.113
%RSD	.60976

#1	18.396
#2	18.526
#3	18.388
#4	18.361
#5	18.625

Method: PRV

Sample Name: s98t002927_d

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 10:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0096	.0088	-.0113	.0356	.0004
Stddev	.0009	.0014	.0121	.0018	.0002
%RSD	9.469	15.88	106.3	5.091	60.55

#1	.0100	.0075	-.0298	.0375	.0006
#2	.0105	.0094	-.0069	.0368	.0001
#3	.0088	.0107	.0033	.0356	.0005
#4	.0084	.0074	-.0095	.0351	.0002
#5	.0101	.0092	-.0138	.0328	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0114	.0055	.0139	.0191
Stddev	.0011	.0006	.0008	.0016
%RSD	9.586	11.59	5.644	8.371

#1	.0125	.0065	.0136	.0208
#2	.0098	.0050	.0143	.0204
#3	.0112	.0050	.0151	.0189
#4	.0111	.0055	.0135	.0181
#5	.0124	.0055	.0131	.0170

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.343
Stddev	.117
%RSD	.63610

#1	18.366
#2	18.481
#3	18.220
#4	18.227
#5	18.423

Analysis Report

HNF-1663 REV. 0

03/07/99 10:41:24

page 1

Method: PRV

Sample Name: s98t002927_a

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 10:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	20.03	.0092	-.0101	3.581	.0002
Stddev	.11	.0007	.0239	.034	.0001
%RSD	.5675	7.521	237.1	.9581	35.29

#1	20.15	.0094	.0270	3.589	.0002
#2	20.06	.0093	-.0299	3.550	.0001
#3	19.92	.0082	-.0317	3.588	.0003
#4	20.11	.0101	-.0035	3.630	.0002
#5	19.90	.0088	-.0123	3.545	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	20.20	.0066	.0148	.0344
Stddev	.33	.0005	.0007	.0007
%RSD	1.615	7.621	4.866	2.070

#1	20.72	.0066	.0144	.0339
#2	19.89	.0062	.0146	.0335
#3	20.10	.0074	.0139	.0343
#4	20.30	.0064	.0154	.0349
#5	20.00	.0063	.0156	.0353

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.261
Stddev	.248
%RSD	1.3593

#1	17.892
#2	18.400
#3	18.416
#4	18.118
#5	18.477

HNF-1663 REV.0

Method: PRV

Sample Name: s98t002927_a

Operator: kjt

Comment: 201 df

Run Time: 03/07/99 10:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0149	.0098	-.0136	19.02	.0004
Stddev	.0010	.0011	.0085	.29	.0003
%RSD	6.925	10.95	62.39	1.503	60.73
#1	.0161	.0110	-.0193	18.76	.0005
#2	.0154	.0094	-.0205	19.32	.0000
#3	.0146	.0094	.0007	19.24	.0004
#4	.0150	.0085	-.0159	19.08	.0005
#5	.0133	.0109	-.0131	18.68	.0007

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0155	.0066	.0138	.0175
Stddev	.0006	.0006	.0004	.0015
%RSD	3.695	8.801	2.884	8.335
#1	.0153	.0063	.0144	.0151
#2	.0156	.0057	.0138	.0177
#3	.0148	.0071	.0138	.0185
#4	.0164	.0070	.0132	.0175
#5	.0153	.0069	.0139	.0188

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	17.730
Stddev	.152
%RSD	.85662

#1	17.937
#2	17.758
#3	17.615
#4	17.551
#5	17.787

Method: PRV Sample Name: ccv

Operator: kjt

Comment: msprv lcs

Run Time: 03/07/99 10:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	18.77	.0022	-.0084	1.279	.0003
Stddev	.14	.0013	.0054	.138	.0002
%RSD	.7283	62.28	64.58	10.76	62.13
#1	18.67	.0010	-.0146	1.085	.0002
#2	18.81	.0015	-.0012	1.207	.0004
#3	18.95	.0043	-.0045	1.299	.0001
#4	18.83	.0025	-.0103	1.373	.0003
#5	18.60	.0015	-.0112	1.433	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.65	.0076	.0132	.1209
Stddev	.24	.0008	.0008	.0029
%RSD	1.211	11.14	5.853	2.414
#1	19.71	.0075	.0143	.1253
#2	19.53	.0072	.0128	.1173
#3	19.98	.0086	.0137	.1207
#4	19.72	.0082	.0124	.1215
#5	19.34	.0064	.0128	.1199

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	18.062
Stddev	.121
%RSD	.66938

#1	18.155
#2	18.009
#3	17.935
#4	17.991
#5	18.223

HNF-1663 REV. 0

Method: PRV

Sample Name: ccv

Operator: kjt

Comment: 19.25 ppb Sn

Run Time: 03/07/99 10:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1938	-.0007	.0157	21.31	.0002
Stddev	.2277	.0007	.0102	.42	.0001
%RSD	117.5	100.5	64.79	1.964	81.65

#1	.5861	-.0013	.0187	20.57	.0001
#2	.1973	-.0002	.0279	21.51	.0000
#3	.0891	-.0013	.0141	21.40	.0003
#4	.0566	.0002	.0179	21.45	.0003
#5	.0400	-.0007	-.0001	21.60	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0129	.0294	.0458	.4288
Stddev	.0017	.0015	.0013	.0109
%RSD	13.15	5.131	2.863	2.543

#1	.0155	.0286	.0460	.4114
#2	.0126	.0278	.0437	.4251
#3	.0132	.0289	.0457	.4360
#4	.0110	.0301	.0465	.4336
#5	.0121	.0317	.0472	.4380

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.771
Stddev	.400
%RSD	2.5367

#1	16.465
#2	15.714
#3	15.580
#4	15.643
#5	15.451

Method: PRV

Sample Name: ccb

Operator: kjt

Comment: 2% nitric

Run Time: 03/07/99 11:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	.0013	-.0108	.0027	.0004
Stddev	.0002	.0015	.0071	.0005	.0003
%RSD	53.05	111.3	65.90	18.38	66.32

#1	-.0002	-.0007	-.0052	.0024	.0007
#2	-.0002	.0026	-.0099	.0021	.0006
#3	-.0003	.0002	-.0040	.0026	.0003
#4	-.0005	.0023	-.0216	.0035	.0002
#5	-.0002	.0023	-.0132	.0029	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0002	.0002	.0001
Stddev	.0002	.0003	.0003	.0006
%RSD	27.13	144.5	141.6	813.6

#1	.0007	-.0004	-.0002	-.0007
#2	.0005	-.0002	.0004	.0002
#3	.0007	-.0003	.0004	-.0005
#4	.0004	-.0002	.0000	.0006
#5	.0008	.0002	.0004	.0008

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	17.747
Stddev	.132
%RSD	.74359

#1	17.671
#2	17.889
#3	17.811
#4	17.557
#5	17.807

LABCORE Completed Worklist Report for Worklist# 27744

Analyst: bjg

Instrument: ICPMS1

Book#: 990226-6

Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSU-QC U235	QC	0.000142	1.50e-04	105.634 %	Recovery
1	ICV	0	EMSU-QC U238	QC	0.020	2.12e-02	106.000 %	Recovery
2	ICB	0	EMSU-QC U233	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMSU-QC U234	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMSU-QC U235	QC	1	<1.20e-5		ug/mL
2	ICB	0	EMSU-QC U236	QC	1	<1.60e-5		ug/mL
2	ICB	0	EMSU-QC U238	QC	1	<1.20e-5		ug/mL
3	STD-PREP	0	EMSU-A1 U238-A1	SOLID	10.0	1.06e+01	106.000 %	Recovery
4	BLNK-PREP	0	EMSU-A1 U233-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMSU-A1 U234-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMSU-A1 U235-A1	SOLID	1	<1.20e-4		ug/g
4	BLNK-PREP	0	EMSU-A1 U236-A1	SOLID	1	<1.60e-4		ug/g
4	BLNK-PREP	0	EMSU-A1 U238-A1	SOLID	1	4.33e-04	4.330e-004	ug/g
5	SAMPLE	S98T002847	0 A	EMSU-A1 U233-A1	SOLID	N/A	5.008e+00	5.009 ug/g
5	SAMPLE	S98T002847	0 A	EMSU-A1 U234-A1	SOLID	N/A	5.008e+00	5.009 ug/g
5	SAMPLE	S98T002847	0 A	EMSU-A1 U235-A1	SOLID	N/A	1.152e+02	5.009 ug/g
5	SAMPLE	S98T002847	0 A	EMSU-A1 U236-A1	SOLID	N/A	6.678e+00	6.677 ug/g
5	SAMPLE	S98T002847	0 A	EMSU-A1 U238-A1	SOLID	N/A	1.768e+04	5.009 ug/g
6	DUP	S98T002847	0 A	EMSU-A1 U233-A1	SOLID	<5.01e0	<4.55e0	RPD
6	DUP	S98T002847	0 A	EMSU-A1 U234-A1	SOLID	<5.01e0	<4.55e0	RPD
6	DUP	S98T002847	0 A	EMSU-A1 U235-A1	SOLID	1.15e+02	1.25e+02	8.333 RPD
6	DUP	S98T002847	0 A	EMSU-A1 U236-A1	SOLID	<6.68e0	<6.07e0	RPD
6	DUP	S98T002847	0 A	EMSU-A1 U238-A1	SOLID	1.77e+04	1.92e+04	8.130 RPD
7	SAMPLE	S98T002926	0 A	EMSU-A1 U233-A1	SOLID	N/A	4.776e-01	0.478 ug/g
7	SAMPLE	S98T002926	0 A	EMSU-A1 U234-A1	SOLID	N/A	4.776e-01	0.478 ug/g
7	SAMPLE	S98T002926	0 A	EMSU-A1 U235-A1	SOLID	N/A	9.871e-01	0.478 ug/g
7	SAMPLE	S98T002926	0 A	EMSU-A1 U236-A1	SOLID	N/A	6.368e-01	0.637 ug/g
7	SAMPLE	S98T002926	0 A	EMSU-A1 U238-A1	SOLID	N/A	1.386e+02	0.478 ug/g
8	DUP	S98T002926	0 A	EMSU-A1 U233-A1	SOLID	<4.78e-1	<4.83e-1	RPD
8	DUP	S98T002926	0 A	EMSU-A1 U234-A1	SOLID	<4.78e-1	<4.83e-1	RPD
8	DUP	S98T002926	0 A	EMSU-A1 U235-A1	SOLID	9.87e-01	7.33e-01	29.535 RPD
8	DUP	S98T002926	0 A	EMSU-A1 U236-A1	SOLID	<6.37e-1	<6.44e-1	RPD
8	DUP	S98T002926	0 A	EMSU-A1 U238-A1	SOLID	1.39e+02	1.13e+02	20.635 RPD
9	SAMPLE	S98T002927	0 A	EMSU-A1 U233-A1	SOLID	N/A	4.726e-01	0.473 ug/g
9	SAMPLE	S98T002927	0 A	EMSU-A1 U234-A1	SOLID	N/A	4.726e-01	0.473 ug/g
9	SAMPLE	S98T002927	0 A	EMSU-A1 U235-A1	SOLID	N/A	6.852e-01	0.473 ug/g
9	SAMPLE	S98T002927	0 A	EMSU-A1 U236-A1	SOLID	N/A	6.301e-01	0.630 ug/g
9	SAMPLE	S98T002927	0 A	EMSU-A1 U238-A1	SOLID	N/A	9.727e+01	0.473 ug/g
10	DUP	S98T002927	0 A	EMSU-A1 U233-A1	SOLID	<4.73e-1	<4.60e-1	RPD
10	DUP	S98T002927	0 A	EMSU-A1 U234-A1	SOLID	<4.73e-1	<4.60e-1	RPD
10	DUP	S98T002927	0 A	EMSU-A1 U235-A1	SOLID	6.85e-01	6.59e-01	3.869 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27744

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
10 DUP	S98T002927	0 A	QMSU-A1	U236-A1	SOLID	<6.30e-1	<6.13e-1	RPD
10 DUP	S98T002927	0 A	QMSU-A1	U236-A1	SOLID	9.79e+01	9.96e+01	2.336 RPD
11 SPK	S98T002927	0 A	QMSU-A1	U235-A1	SOLID	14.22	1.47e+01	103.376 % Recovery
11 SPK	S98T002927	0 A	QMSU-A1	U236-A1	SOLID	2000.00	2.08e+03	104.000 % Recovery
12 SPK-POST	S98T002927	0 A	QMSU-A1	U235-A1	SOLID	0.0001422	1.46e-04	102.672 % Recovery
12 SPK-POST	S98T002927	0 A	QMSU-A1	U236-A1	SOLID	0.020	2.07e-02	103.500 % Recovery
13 CCV	0	0	QMSU-QC	U235	QC	0.000142	1.53e-04	107.746 % Recovery
13 CCV	0	0	QMSU-QC	U236	QC	0.020	2.21e-02	110.500 % Recovery
14 CCB	0	0	QMSU-QC	U233	QC	1	<1.20e-5	ug/mL
14 CCB	0	0	QMSU-QC	U234	QC	1	<1.20e-5	ug/mL
14 CCB	0	0	QMSU-QC	U235	QC	1	<1.20e-5	ug/mL
14 CCB	0	0	QMSU-QC	U236	QC	1	<1.60e-5	ug/mL
14 CCB	0	0	QMSU-QC	U236	QC	1	<1.20e-5	ug/mL

Final page for worklist# 27744

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

3-3-99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27744

Analyst: BTG **Instrument:** ICPMS1 **Book#** 990226-6

Method: LA-506-101 Rev/Mod *A-0*

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	STD-PREP		@MSU-A1	SOLID		
4	BLNK-PREP		@MSU-A1	SOLID		
5	SAMPLE	S98T002847 0 A	@MSU-A1	SOLID	98000437	BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1				
6	DUP	S98T002847 0 A	@MSU-A1	SOLID		
7	SAMPLE	S98T002926 0 A	@MSU-A1	SOLID	98000490	BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1				
8	DUP	S98T002926 0 A	@MSU-A1	SOLID		
9	SAMPLE	S98T002927 0 A	@MSU-A1	SOLID	98000490	BY-105 (2)
		Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1				
10	DUP	S98T002927 0 A	@MSU-A1	SOLID		
11	SPK	S98T002927 0 A	@MSU-A1	SOLID		
12	SPK-POST	S98T002927 0 A	@MSU-A1	SOLID		
13	CCV		@MSU-QC	QC		
14	CCE		@MSU-QC	QC		

Final page for worklist # 27744

Signature [Signature] Date 3-1-88

Signature   **Date**

Data Entry Comments:

Slightly high RPDs on S98T002926 -
be rerun if requested

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

03/01/99 10:23

Page: 2

ws2

LABCORE Data Entry Template for Worklist# 27744

S Type	Sample#	R A Test	Matrix	Group#	Project
	prep STD	.05 - 10ml			201
	prep STD	1ml - 9ml			10
	SFSTW 2847	.05 ml - 10ml - 1ml - 9ml			2010
	2847-D	.05 ml - 10ml - 1ml - 9ml			2010
	SFSTW 2926	.05 ml - 10ml			201
	2926-D	.05 ml - 10ml			
	SFSTW 2927				
	2927-D				
	2927-A				
	2927-S				
	POST SPK - .02ml of 10ppm STD, (20ppm)				
	SCANS = SF2K17D, 982926, 982927				

Data Entry Comments:

Uploaded & Validated by:

Saul H. Peng

03/02/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: bjj

Comment:

Run Time: 03/01/99 10:44 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	211200.
Stddev	1064.
%RSD	.5039

#1	211900.
#2	213300.
#3	211400.
#4	209800.
#5	211500.
#6	209700.
#7	211100.
#8	210300.
#9	211700.
#10	211300.

By 10:15
AKD DEEST
58T00 2847, 2926, 2927
FILE # U0301A.TXT

SCANS = 98284701
982926
982927

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2966 TO 2987

Method: ISOU Sample Name: BLANK

Operator:

Comment:

Run Time: 03/01/99 10:45 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.515	2.340	2.491	2.324	2.999
Stddev	.308	.250	.413	.171	.189
%RSD	12.23	10.68	16.60	7.334	6.296
#1	2.374	2.165	2.124	2.498	2.998
#2	2.726	2.055	2.936	2.516	3.020
#3	2.779	2.655	2.074	2.199	2.904
#4	2.656	2.532	2.905	2.158	2.781
#5	2.041	2.291	2.416	2.250	3.291

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.006
Stddev	.053
%RSD	.43873

#1	12.007
#2	11.921
#3	12.053
#4	12.048
#5	12.001

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	479.2904	2.5153	1.0000000	03/01/99 10:47:11

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.5153
100ppb U	0.711	0.711	0	0	343.29

U [233]

0

(S) IR³⁸⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	479.5374	2.3397	1.0000000	03/01/99 10:47:11

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.3397
100ppb U	0.711	0.711	0	0	343.29

U [234]

0

(S) IR³⁸⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	479.3245	2.4911	1.0000000	03/01/99 10:47:11

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.4911
100ppb U	0.711	0.711	0	0	343.29

U [235]

0

(S) IR³⁸⁰⁰10

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	479.5591	2.3243	1.0000000	03/01/99 10:47:11

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.3243
100ppb U	0.711	0.711	0	0	343.29

U [236]

0

(S) IR³⁸⁰⁰10

Concentration

HNF-1663 REV.0

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [238]	431.7472	2.9987	1.0000000	03/01/99 10:47:11

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	2.9987
100ppb U	99.3	99.3	0	0	42875

U [238]

0

(S) IR 4717010

Concentration

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 03/01/99 10:47 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	655.5	655.5	655.5	655.5	82110.
Stddev	11.0	11.0	11.0	11.0	849.
%RSD	1.678	1.678	1.678	1.678	1.034

#1	643.3	643.3	643.3	643.3	81340.
#2	652.6	652.6	652.6	652.6	81090.
#3	673.5	673.5	673.5	673.5	82910.
#4	654.0	654.0	654.0	654.0	82870.
#5	654.2	654.2	654.2	654.2	82340.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.199
Stddev	.107
%RSD	.87697

#1	12.316
#2	12.065
#3	12.133
#4	12.184
#5	12.296

HNF-1663 REV.0

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	918.4284	2.5153	1.0000000	03/01/99 10:49:14

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.5153
100ppb U	0.711	0.711	0	0	655.52

U [233]

(S) IR⁰₇₃₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	918.6754	2.3397	1.0000000	03/01/99 10:49:14

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.3397
100ppb U	0.711	0.711	0	0	655.52

U [234]

(S) IR⁰₇₃₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	918.4625	2.4911	1.0000000	03/01/99 10:49:14

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.4911
100ppb U	0.711	0.711	0	0	655.52

U [235]

(S) IR⁰₇₃₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	918.6971	2.3243	1.0000000	03/01/99 10:49:14

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.3243
100ppb U	0.711	0.711	0	0	655.52

U [236]

(S) IR⁰₇₃₀₀₁₀

Concentration

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	826.8541	2.9987	1.0000000	03/01/99 10:49:14

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	2.9987
100ppb U	99.3	99.3	0	0	82110

U [238]

0

(S) IR 90330110

Concentration

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/01/99 10:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	.0001	-.0003	.0002	-.0003
Stddev	.0005	.0005	.0002	.0003	.0002
%RSD	254.0	978.0	89.00	203.6	89.74

#1	-.0001	.0005	-.0004	.0006	-.0005
#2	-.0007	-.0003	.0000	-.0003	.0001
#3	-.0006	-.0006	-.0005	.0004	-.0002
#4	.0002	.0005	-.0004	.0001	-.0002
#5	.0003	.0002	-.0001	.0001	-.0004

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.391
Stddev	.109
%RSD	.87988

#1	12.446
#2	12.532
#3	12.242
#4	12.395
#5	12.341

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/01/99 10:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	.0015	.1504	.0026	21.24
Stddev	.0002	.0003	.0013	.0003	.30
%RSD	304.1	19.27	.8412	10.91	1.421
#1	.0000	.0015	.1486	.0029	21.44
#2	.0000	.0017	.1503	.0028	21.25
#3	.0002	.0017	.1503	.0022	20.74
#4	-.0003	.0018	.1511	.0024	21.23
#5	-.0003	.0010	.1520	.0025	21.51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.541
Stddev	.104
%RSD	.82964

#1	12.445
#2	12.416
#3	12.577
#4	12.626
#5	12.640

Method: ISOÜ

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 11:07 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	.0001	-.0004	.0001	-.0001
Stddev	.0004	.0004	.0003	.0004	.0005
%RSD	80.10	255.2	95.58	466.5	1021.
#1	-.0010	.0003	-.0006	-.0003	-.0008
#2	-.0010	.0001	-.0008	.0000	.0000
#3	-.0002	.0000	.0001	.0000	.0000
#4	.0000	-.0003	-.0004	.0000	-.0001
#5	-.0006	.0006	-.0001	.0007	.0007

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	12.614
Stddev	.138
%RSD	1.0965

#1	12.558
#2	12.493
#3	12.597
#4	12.852
#5	12.573

Method: ISOU Sample Name: prepstd

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0001	.0031	.3758	.0052	52.96
Stddev	.0002	.0006	.0055	.0006	.55
%RSD	185.2	20.91	1.464	10.91	1.046
#1	-.0001	.0030	.3766	.0052	53.47
#2	.0002	.0034	.3786	.0045	52.31
#3	-.0003	.0040	.3776	.0054	53.56
#4	-.0002	.0025	.3801	.0061	52.91
#5	-.0001	.0025	.3663	.0051	52.53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	13.032
Stddev	.123
%RSD	.94451

#1	12.946
#2	13.093
#3	12.914
#4	12.992
#5	13.216

Method: ISOU Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/01/99 11:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	-.0002	-.0003	-.0002	.0433
Stddev	.0004	.0002	.0002	.0002	.0023
%RSD	120.9	94.82	80.79	109.4	5.261

#1	-.0003	-.0005	.0000	-.0003	.0452
#2	.0003	-.0002	-.0002	-.0002	.0441
#3	-.0002	-.0001	-.0002	-.0004	.0410
#4	-.0006	-.0002	-.0004	.0002	.0454
#5	-.0005	.0000	-.0005	-.0002	.0407

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.234
Stddev	.258
%RSD	1.9527

#1	13.069
#2	13.208
#3	13.254
#4	12.984
#5	13.654

Method: ISOU

Sample Name: s98t002847

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 11:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0025	.2760	.0068	42.36
Stddev	.0003	.0004	.0072	.0004	.77
%RSD	205.5	14.35	2.593	6.548	1.812
#1	.0002	.0021	.2807	.0073	42.80
#2	-.0002	.0024	.2830	.0070	43.12
#3	.0004	.0023	.2683	.0066	41.91
#4	-.0001	.0031	.2683	.0061	41.25
#5	.0005	.0025	.2799	.0068	42.74

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.159
Stddev	.219
%RSD	1.6616

#1	12.900
#2	12.967
#3	13.346
#4	13.385
#5	13.199

Method: ISOU

Sample Name: s98t002847_d

Operator: bjg

Comment: 2010 df

Run Time: 03/01/99 11:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	.0030	.3298	.0079	50.59
Stddev	.0003	.0005	.0072	.0007	1.00
%RSD	186.8	15.02	2.182	8.538	1.975
#1	-.0003	.0026	.3195	.0070	48.98
#2	-.0003	.0037	.3369	.0084	51.57
#3	-.0003	.0028	.3261	.0082	50.44
#4	-.0003	.0032	.3309	.0084	50.79
#5	.0004	.0028	.3358	.0073	51.20

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.124
Stddev	.321
%RSD	2.4447

#1	13.618
#2	12.761
#3	13.211
#4	12.967
#5	13.063

Method: ISOU   Sample Name: s98t002926   Operator: bjg
Comment: 201 df
Run Time: 03/01/99 11:33 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0021	-.0001	.0248	.0007	3.482
Stddev	.0005	.0004	.0007	.0003	.080
%RSD	22.06	675.1	2.905	48.79	2.295

#1	.0017	-.0002	.0247	.0002	3.402
#2	.0021	-.0006	.0246	.0009	3.530
#3	.0016	.0002	.0256	.0005	3.458
#4	.0024	.0005	.0255	.0006	3.595
#5	.0027	-.0003	.0238	.0010	3.423

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.427
Stddev	.249
%RSD	1.8574

#1	13.739
#2	13.186
#3	13.537
#4	13.156
#5	13.516

Method: ISOU

Sample Name: s98t002926_d

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0019	-.0006	.0182	.0003	2.806
Stddev	.0006	.0002	.0014	.0002	.165
%RSD	33.71	44.87	7.459	55.72	5.870
#1	.0008	-.0009	.0161	.0005	2.522
#2	.0025	-.0003	.0192	.0003	2.859
#3	.0021	-.0004	.0195	.0000	2.931
#4	.0019	-.0005	.0179	.0005	2.815
#5	.0020	-.0006	.0181	.0004	2.901

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.317
Stddev	.763
%RSD	5.3325

#1	15.636
#2	14.013
#3	13.819
#4	14.299
#5	13.816

Method: ISOU

Sample Name: s98t002927

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	-.0002	.0174	.0002	2.470
Stddev	.0004	.0004	.0005	.0003	.095
%RSD	38.01	167.9	3.070	153.5	3.840
#1	.0012	-.0007	.0172	-.0001	2.313
#2	.0009	.0003	.0168	.0000	2.513
#3	.0006	-.0002	.0181	.0005	2.492
#4	.0013	-.0004	.0178	.0003	2.467
#5	.0006	-.0001	.0171	.0001	2.565

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.014
Stddev	.334
%RSD	2.3867

#1	14.600
#2	13.818
#3	13.962
#4	13.898
#5	13.793

Method: ISOU

Sample Name: s98t002927_2-5 ³³⁸⁵

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0009	.0034	.3975	.0061	56.47
Stddev	.0003	.0007	.0046	.0004	.63
%RSD	32.78	19.64	1.156	6.407	1.113
#1	.0013	.0043	.3906	.0063	55.58
#2	.0012	.0035	.4019	.0057	56.11
#3	.0008	.0025	.4014	.0064	56.61
#4	.0008	.0034	.3979	.0058	56.94
#5	.0006	.0032	.3958	.0066	57.12

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.788
Stddev	.114
%RSD	.82607

#1	13.937
#2	13.783
#3	13.656
#4	13.705
#5	13.861

Method: ISOU

Sample Name: s98t002927_a

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0010	.0009	.1631	.0026	23.17
Stddev	.0002	.0003	.0042	.0005	.46
%RSD	24.56	30.96	2.552	19.91	1.972
#1	.0007	.0012	.1582	.0023	23.13
#2	.0009	.0007	.1593	.0018	22.89
#3	.0013	.0011	.1674	.0029	23.84
#4	.0012	.0006	.1664	.0030	22.65
#5	.0009	.0007	.1643	.0028	23.35

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	13.883
Stddev	.203
%RSD	1.4653

#1	14.058
#2	14.085
#3	13.597
#4	13.773
#5	13.903

Method: ISOU

Sample Name: s98t002927 *sd 6/13/99*

Operator: bjg

Comment: 201 df

Run Time: 03/01/99 11:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0012	-.0001	.0172	.0002	2.600
Stddev	.0006	.0004	.0014	.0002	.174
%RSD	50.19	397.7	8.128	110.9	6.678
#1	.0002	-.0003	.0155	.0003	2.293
#2	.0014	-.0005	.0184	.0003	2.690
#3	.0016	.0005	.0174	.0002	2.710
#4	.0018	.0002	.0187	-.0002	2.662
#5	.0012	-.0005	.0161	.0004	2.644

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.474
Stddev	.944
%RSD	6.5210

#1	16.132
#2	13.824
#3	13.964
#4	14.167
#5	14.283

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/01/99 11:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	.0009	.1528	.0017	22.07
Stddev	.0003	.0003	.0034	.0003	.32
%RSD	67.35	38.21	2.226	17.50	1.461
#1	-.0007	.0006	.1471	.0015	21.63
#2	-.0003	.0012	.1539	.0012	22.13
#3	-.0006	.0009	.1557	.0018	22.52
#4	-.0006	.0012	.1522	.0020	21.96
#5	.0000	.0004	.1548	.0018	22.11

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.902
Stddev	.269
%RSD	1.9359

#1	14.308
#2	13.929
#3	13.557
#4	13.832
#5	13.883

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/01/99 11:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	-.0004	-.0003	-.0003	.0003
Stddev	.0002	.0003	.0004	.0003	.0005
%RSD	34.10	71.17	137.5	104.8	144.5

#1	-.0009	-.0004	-.0007	.0000	.0002
#2	-.0006	-.0004	.0003	.0001	.0001
#3	-.0003	-.0006	-.0004	-.0004	.0008
#4	-.0008	-.0005	-.0001	-.0007	.0008
#5	-.0006	.0001	-.0006	-.0006	-.0003

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	13.831
Stddev	.130
%RSD	.94346

#1	13.823
#2	14.007
#3	13.639
#4	13.851
#5	13.837

LABCORE Completed Worklist Report for Worklist# 27745

Analyst: kjt

Instrument: ICPMS1

Book#: 990223-1

Method: LA-506-101 Rev/Mod A0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	QMSU-QC U235	QC	0.000142	1.57e-04	110.563 % Recovery	
1 ICV	0	QMSU-QC U238	QC	0.020	2.01e-02	100.500 % Recovery	
2 ICB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2 ICB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2 ICB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2 ICB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2 ICB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3 STD-PREP	0	QMSU-A1 U238-A1	SOLID	10	9.15e+00	91.500 % Recovery	
4 BLNK-PREP	0	QMSU-A1 U233-A1	SOLID	1	<1.32e-4	ug/g	
4 BLNK-PREP	0	QMSU-A1 U234-A1	SOLID	1	<1.32e-4	ug/g	
4 BLNK-PREP	0	QMSU-A1 U235-A1	SOLID	1	<1.32e-4	ug/g	
4 BLNK-PREP	0	QMSU-A1 U236-A1	SOLID	1	<1.76e-4	ug/g	
4 BLNK-PREP	0	QMSU-A1 U238-A1	SOLID	1	<1.32e-4	ug/g	
5 SAMPLE	S98T002607	0 A	QMSU-A1 U233-A1	SOLID	N/A <	4.683e-01	0.468 ug/g
5 SAMPLE	S98T002607	0 A	QMSU-A1 U234-A1	SOLID	N/A <	4.683e-01	0.468 ug/g
5 SAMPLE	S98T002607	0 A	QMSU-A1 U235-A1	SOLID	N/A	5.308e-01	0.468 ug/g
5 SAMPLE	S98T002607	0 A	QMSU-A1 U236-A1	SOLID	N/A <	6.245e-01	0.625 ug/g
5 SAMPLE	S98T002607	0 A	QMSU-A1 U238-A1	SOLID	N/A	4.574e+01	0.468 ug/g
6 DUP	S98T002607	0 A	QMSU-A1 U233-A1	SOLID	<4.68e-1	<4.82e-1	RPD
6 DUP	S98T002607	0 A	QMSU-A1 U234-A1	SOLID	<4.68e-1	<4.82e-1	RPD
6 DUP	S98T002607	0 A	QMSU-A1 U235-A1	SOLID	5.31e-01	5.54e-01	4.240 RPD
6 DUP	S98T002607	0 A	QMSU-A1 U236-A1	SOLID	<6.25e-1	<6.42e-1	RPD
6 DUP	S98T002607	0 A	QMSU-A1 U238-A1	SOLID	4.57e+01	5.00e+01	8.986 RPD
7 SAMPLE	S98T002608	0 A	QMSU-A1 U233-A1	SOLID	N/A <	4.818e+01	48.170 ug/g
7 SAMPLE	S98T002608	0 A	QMSU-A1 U234-A1	SOLID	N/A <	4.818e+01	48.170 ug/g
7 SAMPLE	S98T002608	0 A	QMSU-A1 U235-A1	SOLID	N/A	2.156e+02	48.170 ug/g
7 SAMPLE	S98T002608	0 A	QMSU-A1 U236-A1	SOLID	N/A <	6.424e+01	64.250 ug/g
7 SAMPLE	S98T002608	0 A	QMSU-A1 U238-A1	SOLID	N/A	3.075e+04	48.170 ug/g
8 DUP	S98T002608	0 A	QMSU-A1 U233-A1	SOLID	<4.82e1	<4.71e1	RPD
8 DUP	S98T002608	0 A	QMSU-A1 U234-A1	SOLID	<4.82e1	<4.71e1	RPD
8 DUP	S98T002608	0 A	QMSU-A1 U235-A1	SOLID	2.16e+02	2.54e+02	16.170 RPD
8 DUP	S98T002608	0 A	QMSU-A1 U236-A1	SOLID	<6.42e1	<6.28e1	RPD
8 DUP	S98T002608	0 A	QMSU-A1 U238-A1	SOLID	3.08e+04	3.45e+04	11.332 RPD
9 SFK	S98T002608	0 A	QMSU-A1 U235-A1	SOLID	24.44	2.76e+01	112.930 % Recovery
9 SFK	S98T002608	0 A	QMSU-A1 U238-A1	SOLID	2000.0	1.81e+03	90.500 % Recovery
10 SFK-POST	S98T002608	0 A	QMSU-A1 U235-A1	SOLID	0.0001422	1.31e-04	92.124 % Recovery
10 SFK-POST	S98T002608	0 A	QMSU-A1 U238-A1	SOLID	0.020	1.96e-02	98.000 % Recovery
11 SAMPLE	S98T002609	0 A	QMSU-A1 U233-A1	SOLID	N/A <	4.955e+01	49.550 ug/g
11 SAMPLE	S98T002609	0 A	QMSU-A1 U234-A1	SOLID	N/A <	4.955e+01	49.550 ug/g
11 SAMPLE	S98T002609	0 A	QMSU-A1 U235-A1	SOLID	N/A	3.188e+02	49.550 ug/g
11 SAMPLE	S98T002609	0 A	QMSU-A1 U236-A1	SOLID	N/A <	6.607e+01	66.080 ug/g

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27745

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11 SAMPLE	S98T002609	0 A	QMSU-A1 U238-A1	SOLID	N/A	4.419e+04	49.550	ug/g
12 DUP	S98T002609	0 A	QMSU-A1 U233-A1	SOLID	<4.95e1	<4.66e1		RPD
12 DUP	S98T002609	0 A	QMSU-A1 U234-A1	SOLID	<4.95e1	<4.66e1		RPD
12 DUP	S98T002609	0 A	QMSU-A1 U235-A1	SOLID	3.19e+02	2.87e+02	10.561	RPD
12 DUP	S98T002609	0 A	QMSU-A1 U236-A1	SOLID	<6.61e1	<6.21e1		RPD
12 DUP	S98T002609	0 A	QMSU-A1 U238-A1	SOLID	4.42e+04	3.89e+04	12.756	RPD
13 CCV		0	QMSU-QC U235	QC	0.000142	1.48e-04	104.225	% Recovery
13 CCV		0	QMSU-QC U238	QC	0.020	1.93e-02	96.500	% Recovery
14 CCB		0	QMSU-QC U233	QC	1	<1.20e-5		ug/mL
14 CCB		0	QMSU-QC U234	QC	1	<1.20e-5		ug/mL
14 CCB		0	QMSU-QC U235	QC	1	<1.20e-5		ug/mL
14 CCB		0	QMSU-QC U236	QC	1	<1.60e-5		ug/mL
14 CCB		0	QMSU-QC U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 27745

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date _____

12/31/98 09:53
A-0004-1

HNH-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 27745

Analyst: KSP Instrument: ICPMS1 Poems Book# 990223-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		@MSU-QC	QC		
2	ICB		@MSU-QC	QC		
3	STD-PREP		@MSU-A1	SOLID		
4	BINK-PREP		@MSU-A1	SOLID		
5	SAMPLE	S98T002607 0 A	@MSU-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
6	DUP	S98T002607 0 A	@MSU-A1	SOLID		
7	SAMPLE	S98T002608 0 A	@MSU-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
8	DUP	S98T002608 0 A	@MSU-A1	SOLID		
9	SPK	S98T002608 0 A	@MSU-A1	SOLID		
10	SPK-POST	S98T002608 0 A	@MSU-A1	SOLID		
11	SAMPLE	S98T002609 0 A	@MSU-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1					
12	DUP	S98T002609 0 A	@MSU-A1	SOLID		
13	CCV		@MSU-QC	QC		
14	CCB		@MSU-QC	QC		

Final page for worklist # 27745

Ks Horndike 2-25-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

file v0225a.txt

Upgraded & Validated by
Sam H. Pang

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2/26/99

2990

12/31/98 09:53
A-0004-1

LABCORE Data Entry Template for Worklist# 27745

S Type	Sample#	R A Test	Matrix	Group#	Project
				201	
	Std. Prep			207	.05-10
	Prep-BIK			11	1-16
	S98T002607			201	.05-10
	2607-d				
	2607-S				
	~~2607-a~~			20301	~~.05-10~~ - 1-10
Change in	-				
DF					
	2608				
	2608-d				
	2608-S				
	2608-a				
	2609				
	2609-d				

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

2991

Method: MS_TUNE Sample Name: ms-tune

Operator: kjt

Comment:

Run Time: 02/25/99 10:45 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	384600.
Stddev	5029.
%RSD	1.308

#1	385400.
#2	377200.
#3	387900.
#4	374600.
#5	386600.
#6	384300.
#7	387800.
#8	391500.
#9	384800.
#10	385700.

SP
2/25/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 2992 TO 3015

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 02/25/99 10:47 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.191	.9793	1.104	1.130	1.332
Stddev	.089	.0948	.198	.225	.077
%RSD	7.495	9.679	17.94	19.91	5.756

#1	1.291	1.096	1.202	1.079	1.344
#2	1.133	.9708	.9348	1.528	1.384
#3	1.282	1.040	1.245	.9846	1.412
#4	1.099	.8492	.8492	1.016	1.216
#5	1.149	.9400	1.288	1.044	1.306

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	28.351
Stddev	1.152
%RSD	4.0642

#1	28.279
#2	27.813
#3	26.913
#4	30.027
#5	28.723

Method : ISOU

Standardzn Report

02/25/99 10:49:19

page 1

HNF-1663 REV.0

El Name	Slope	Y-int	Correlation	Date Stdized
J_[233]	392.7892	1.1906	1.0000000	02/25/99 10:49:07

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1906
100ppb U	0.711	0.711	0	0	280.46

U [233]

0

(S) IR³¹⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	393.0864	0.9793	1.0000000	02/25/99 10:49:07

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	0.97933
100ppb U	0.711	0.711	0	0	280.46

U [234]

0

(S) IR³¹⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	392.9112	1.1039	1.0000000	02/25/99 10:49:07

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1039
100ppb U	0.711	0.711	0	0	280.46

U [235]

0

(S) IR³¹⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	392.8740	1.1303	1.0000000	02/25/99 10:49:07

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1303
100ppb U	0.711	0.711	0	0	280.46

U [236]

0

(S) IR³¹⁰⁰¹⁰

Concentration

ethod : ISOU

Standardzn Report

02/25/99 10:49:20

page 2

HNF-1663 REV. 0

Sl Name	Slope	Y-int	Correlation	Date Stdized
J_[238]	361.0275	1.3322	1.0000000	02/25/99 10:49:07

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S)IR
Blank	0	0	0	0	1.3322
100ppb U	99.3	99.3	0	0	35851

U [238]

0
39440m10
(S)IR

Concentration

analysis Report

02/25/99 10:56:55

page 1
HNF-1663 REV.0

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 02/25/99 10:55 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	292.6	292.6	292.6	292.6	39170.
Stddev	19.4	19.4	19.4	19.4	1470.
%RSD	6.634	6.634	6.634	6.634	3.752

#1	288.6	288.6	288.6	288.6	40020.
#2	278.7	278.7	278.7	278.7	38460.
#3	278.7	278.7	278.7	278.7	37710.
#4	325.8	325.8	325.8	325.8	41310.
#5	291.0	291.0	291.0	291.0	38350.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	16.187
Stddev	.746
%RSD	4.6087

#1	16.418
#2	16.547
#3	16.506
#4	14.858
#5	16.606

Method : ISOU

Standardzn Report

02/25/99 10:57:03

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
U [233]	409.7896	1.1906	1.0000000	02/25/99 10:56:53

HNF-1663 REV. 0

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1906
100ppb U	0.711	0.711	0	0	292.55

U [233]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U [234]	410.0868	0.9793	1.0000000	02/25/99 10:56:53

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	0.97933
100ppb U	0.711	0.711	0	0	292.55

U [234]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U [235]	409.9116	1.1039	1.0000000	02/25/99 10:56:53

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1039
100ppb U	0.711	0.711	0	0	292.55

U [235]

0
(S) IR³³⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U [236]	409.8744	1.1303	1.0000000	02/25/99 10:56:53

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.1303
100ppb U	0.711	0.711	0	0	292.55

U [236]

0
(S) IR³³⁰⁰¹⁰

Concentration

Method : ISOU

Standardzn Report

02/25/99 10:57:03

page 2

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U [238]	394.4507	1.3322	1.0000000	02/25/99 10:56:53

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	1.3322
100ppb U	99.3	99.3	0	0	39170

U [238]

0

(S) IR 43990110

Concentration

Method: ISOU Sample Name: blank

Operator: kjt

Comment: background

Run Time: 02/25/99 11:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0038	.0044	.0041	.0043	.0058
Stddev	.0003	.0006	.0007	.0006	.0013
%RSD	6.815	13.89	17.53	13.24	22.16

#1	.0037	.0046	.0048	.0045	.0069
#2	.0039	.0044	.0037	.0044	.0041
#3	.0041	.0038	.0030	.0041	.0057
#4	.0038	.0041	.0044	.0034	.0051
#5	.0034	.0054	.0044	.0050	.0072

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	16.738
Stddev	.895
%RSD	5.3501

#1	16.775
#2	17.808
#3	17.004
#4	16.775
#5	15.328

HNF-1663 REV. 0

Method: ISOU Sample Name: blank

Operator: kjt

Comment: background

Run Time: 02/25/99 11:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0052	.0042	.0044	.0051	.0049
Stddev	.0005	.0005	.0013	.0009	.0008
%RSD	9.321	12.79	29.25	17.87	17.22

#1	.0045	.0044	.0043	.0057	.0038
#2	.0056	.0034	.0036	.0059	.0043
#3	.0052	.0047	.0065	.0053	.0060
#4	.0051	.0041	.0045	.0049	.0054
#5	.0057	.0045	.0031	.0036	.0051

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.829
Stddev	.472
%RSD	3.1861

#1	15.182
#2	14.437
#3	14.536
#4	14.508
#5	15.481

Method: ISOV Sample Name: ICV

Operator: kjt

Comment: iso-u lcs

Run Time: 02/25/99 11:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0054	.0059	.1567	.0084	20.09
Stddev	.0007	.0010	.0098	.0011	1.47
%RSD	12.20	17.01	6.274	12.77	7.335

#1	.0053	.0057	.1548	.0094	18.30
#2	.0062	.0047	.1443	.0093	20.66
#3	.0058	.0073	.1613	.0071	21.92
#4	.0046	.0054	.1525	.0086	18.88
#5	.0049	.0065	.1705	.0074	20.67

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	14.089
Stddev	.715
%RSD	5.0763

#1	14.534
#2	14.704
#3	13.027
#4	14.505
#5	13.676

Method: ISOU Sample Name: ICV

Operator: kjt

Comment: iso-u lcs

Run Time: 02/25/99 11:16 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0040	.0053	.1401	.0070	19.12
Stddev	.0014	.0005	.0111	.0014	1.97
%RSD	34.60	10.09	7.907	20.32	10.32

#1	.0036	.0057	.1554	.0075	18.85
#2	.0042	.0060	.1458	.0082	20.85
#3	.0033	.0054	.1355	.0060	19.58
#4	.0063	.0051	.1375	.0082	20.45
#5	.0027	.0046	.1261	.0051	15.88

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	15.413
Stddev	1.060
%RSD	6.8773

#1	15.566
#2	14.617
#3	15.539
#4	14.313
#5	17.029

Method: ISOU

Sample Name: icb

Operator: kjt

Comment: 2% nitric

Run Time: 02/25/99 11:23 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0042	.0051	.0040	.0046	.0049
Stddev	.0008	.0006	.0008	.0008	.0011
%RSD	18.14	11.38	19.71	17.34	21.94
#1	.0048	.0052	.0047	.0043	.0050
#2	.0035	.0047	.0044	.0044	.0063
#3	.0051	.0043	.0033	.0058	.0035
#4	.0044	.0057	.0046	.0050	.0043
#5	.0034	.0055	.0030	.0037	.0056

Int. Std.

Ir[193]

Line 193/pulse

Units Cts/S

Avg 15.247

Stddev 1.140

%RSD 7.4766

#1	14.274
#2	16.575
#3	15.272
#4	13.958
#5	16.156

HNF-1663 REV. 0

Method: ISOU

Sample Name: std-prep

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 11:25 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0037	.0076	.3453	.0109	45.54
Stddev	.0007	.0009	.0220	.0015	3.15
%RSD	19.43	12.34	6.358	14.09	5.907
#1	.0026	.0089	.3635	.0106	43.60
#2	.0038	.0068	.3380	.0102	48.77
#3	.0034	.0068	.3346	.0089	41.04
#4	.0045	.0072	.3720	.0125	47.24
#5	.0041	.0081	.3185	.0124	47.06

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	15.586
Stddev	.512
%RSD	3.2881

#1	14.754
#2	16.031
#3	15.510
#4	15.973
#5	15.661

HNF-1663 REV. 0

Method: ISOU

Sample Name: prep-blk

Operator: kjt

Comment: 11 df

Run Time: 02/25/99 11:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0039	.0045	.0043	.0038	.0061
Stddev	.0022	.0016	.0021	.0020	.0047
%RSD	57.14	35.30	48.81	52.35	77.80
#1	.0034	.0038	.0042	.0025	.0002
#2	.0005	.0022	.0012	.0009	.0025
#3	.0065	.0062	.0062	.0050	.0092
#4	.0050	.0043	.0038	.0055	.0066
#5	.0043	.0057	.0064	.0052	.0117

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	16.887
Stddev	5.700
%RSD	33.757

#1	19.399
#2	25.793
#3	12.896
#4	13.819
#5	12.526

Method: ISOU

Sample Name: s98t002607

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 11:34 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0042	.0036	.0136	.0036	1.172
Stddev	.0009	.0005	.0010	.0008	.038
%RSD	22.65	14.48	7.269	22.63	3.265
#1	.0045	.0028	.0138	.0039	1.113
#2	.0032	.0040	.0136	.0047	1.161
#3	.0041	.0035	.0142	.0038	1.198
#4	.0056	.0040	.0145	.0027	1.173
#5	.0035	.0036	.0120	.0029	1.212

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	18.999
Stddev	.793
%RSD	4.1726

#1	19.435
#2	19.189
#3	17.639
#4	19.076
#5	19.656

Method: ISOU

Sample Name: s98t002607_d

Operator: kjt

Comment: 201 df

Run Time: 02/25/99 11:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0051	.0037	.0138	.0046	1.245
Stddev	.0008	.0007	.0022	.0006	.062
%RSD	16.67	18.85	16.03	13.33	5.002
#1	.0058	.0046	.0171	.0050	1.272
#2	.0054	.0037	.0149	.0051	1.336
#3	.0047	.0031	.0124	.0050	1.229
#4	.0038	.0042	.0125	.0042	1.171
#5	.0056	.0030	.0119	.0037	1.218

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	18.072
Stddev	.946
%RSD	5.2373

#1	17.354
#2	17.104
#3	17.836
#4	19.374
#5	18.689

Method: ISOU

Sample Name: s98t002608

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99

11:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0031	.0041	.0537	.0048	7.658
Stddev	.0006	.0007	.0057	.0010	.730
%RSD	18.21	17.07	10.69	21.83	9.531
#1	.0025	.0045	.0512	.0049	7.930
#2	.0029	.0032	.0521	.0038	7.165
#3	.0039	.0038	.0638	.0062	8.741
#4	.0029	.0038	.0521	.0052	6.859
#5	.0033	.0050	.0493	.0038	7.595

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	20.708
Stddev	1.876
%RSD	9.0581

#1	19.787
#2	21.539
#3	18.335
#4	23.331
#5	20.549

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002608_d

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99

11:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0033	.0050	.0646	.0050	8.779
Stddev	.0008	.0013	.0058	.0007	.987
%RSD	25.52	25.40	8.925	13.37	11.24
#1	.0030	.0052	.0634	.0046	8.868
#2	.0043	.0072	.0736	.0057	10.36
#3	.0022	.0045	.0665	.0054	8.779
#4	.0032	.0040	.0598	.0052	8.020
#5	.0039	.0043	.0597	.0040	7.872

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	18.602
Stddev	1.802
%RSD	9.6895

#1	17.693
#2	15.879
#3	19.544
#4	20.296
#5	19.600

HNF-1663 REV.0

Method: ISOU

Sample Name: s98t002608_s

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99 12:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0031	.0040	.0635	.0045	8.500
Stddev	.0010	.0008	.0107	.0011	.905
%RSD	32.84	20.38	16.89	24.58	10.64
#1	.0040	.0048	.0634	.0050	8.740
#2	.0030	.0039	.0651	.0049	8.076
#3	.0043	.0049	.0789	.0048	9.296
#4	.0018	.0032	.0487	.0025	7.142
#5	.0026	.0033	.0616	.0052	9.247

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.468
Stddev	2.288
%RSD	11.754

#1	17.523
#2	19.192
#3	17.733
#4	23.214
#5	19.678

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002608_a

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99 12:03 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0027	.0051	.1851	.0071	27.22
Stddev	.0004	.0009	.0121	.0011	1.53
%RSD	14.62	18.06	6.532	14.94	5.610
#1	.0029	.0044	.1940	.0069	27.95
#2	.0023	.0039	.1693	.0060	25.93
#3	.0025	.0055	.1814	.0074	28.70
#4	.0025	.0063	.2000	.0088	28.25
#5	.0033	.0055	.1807	.0065	25.25

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.586
Stddev	.363
%RSD	1.8514

#1	19.845
#2	19.842
#3	19.389
#4	19.037
#5	19.818

Method: ISOU

Sample Name: s98t002609

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99 12:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0032	.0046	.0772	.0057	10.70
Stddev	.0008	.0009	.0086	.0005	.66
%RSD	23.72	20.34	11.20	8.661	6.141
#1	.0031	.0042	.0693	.0060	10.59
#2	.0032	.0052	.0920	.0055	11.51
#3	.0024	.0036	.0738	.0050	10.86
#4	.0028	.0041	.0751	.0055	9.699
#5	.0044	.0059	.0758	.0062	10.86

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	18.849
Stddev	1.194
%RSD	6.3355

#1	18.435
#2	18.860
#3	19.681
#4	20.176
#5	17.094

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002609_d

Operator: kjt

Comment: 20301 df

Run Time: 02/25/99 12:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0035	.0043	.0739	.0057	10.01
Stddev	.0007	.0009	.0064	.0007	.94
%RSD	19.35	21.01	8.666	12.40	9.389
#1	.0040	.0042	.0813	.0064	10.76
#2	.0028	.0034	.0725	.0056	9.794
#3	.0043	.0058	.0795	.0050	11.17
#4	.0028	.0042	.0704	.0050	8.908
#5	.0038	.0041	.0659	.0064	9.404

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	19.177
Stddev	1.193
%RSD	6.2220

#1	17.772
#2	19.305
#3	18.240
#4	20.716
#5	19.854

HNF-1663 REV. 0

Method: ISOU

Sample Name: ccv

Operator: kjt

Comment: iso-u lcs

Run Time: 02/25/99 12:17 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0035	.0041	.1484	.0058	19.31
Stddev	.0006	.0008	.0108	.0006	1.32
%RSD	18.63	18.34	7.270	9.509	6.852
#1	.0038	.0038	.1339	.0050	17.36
#2	.0033	.0036	.1608	.0058	20.34
#3	.0044	.0052	.1552	.0061	19.39
#4	.0031	.0045	.1508	.0058	20.66
#5	.0027	.0033	.1414	.0065	18.78

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	19.472
Stddev	1.317
%RSD	6.7630

#1	21.720
#2	18.282
#3	18.971
#4	19.022
#5	19.365

HNF-1663 REV. 0

Method: ISOU Sample Name: ccb

Operator: kjt

Comment: 2%hno3

Run Time: 02/25/99 12:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0030	.0036	.0025	.0029	.0036
Stddev	.0011	.0011	.0011	.0011	.0011
%RSD	34.91	30.04	42.52	37.38	31.30

#1	.0013	.0018	.0009	.0012	.0023
#2	.0028	.0036	.0027	.0028	.0038
#3	.0040	.0046	.0035	.0037	.0035
#4	.0038	.0040	.0021	.0029	.0031
#5	.0033	.0043	.0034	.0040	.0054

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	21.666
Stddev	4.027
%RSD	18.588

#1	28.607
#2	19.930
#3	20.193
#4	21.305
#5	18.293

LABCORE Completed Worklist Report for Worklist# 27746

Analyst: bjg

Instrument: ICPMS1

Book#: 48 B57
49 B57

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSPR-QC TC99	QC	0.02050	1.92e-02	93.659 % Recovery	
1	ICV	0	QMSPR-QC SN117	QC	0.018560	2.14e-02	115.302 % Recovery	
1	ICV	0	QMSPR-QC CS133	QC	0.02000	1.98e-02	99.000 % Recovery	
2	ICB	0	QMSPR-QC TC99	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC SN117	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC SN126	QC	1	<1.50e-4	ug/mL	
2	ICB	0	QMSPR-QC CS133	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC CS135	QC	1	<2.00e-6	ug/mL	
2	ICB	0	QMSPR-QC CS137	QC	1	<2.00e-6	ug/mL	
3	BLNK-PREP	0	QMSPR-A1 TC99-A1	SOLID	1	4.91e-04	4.910e-004 ug/g	
3	BLNK-PREP	0	QMSPR-A1 SN117-A1	SOLID	1	7.64e-03	0.008 ug/g	
3	BLNK-PREP	0	QMSPR-A1 SN126-A1	SOLID	1	<1.50e-3	ug/g	
3	BLNK-PREP	0	QMSPR-A1 CS133-A1	SOLID	1	<2.00e-5	ug/g	
3	BLNK-PREP	0	QMSPR-A1 CS135-A1	SOLID	1	6.50e-05	6.500e-005 ug/g	
3	BLNK-PREP	0	QMSPR-A1 CS137-A1	SOLID	1	1.12e-04	1.120e-004 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 TC99-A1	SOLID	N/A	1.452e+00	7.98e-002 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 SN117-A1	SOLID	N/A	5.982e+00	5.982 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 SN126-A1	SOLID	N/A	5.982e+00	5.982 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 CS133-A1	SOLID	N/A	3.035e+00	7.98e-002 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 CS135-A1	SOLID	N/A	4.479e+00	7.98e-002 ug/g	
4	SAMPLE	S98T002713	0 A QMSPR-A1 CS137-A1	SOLID	N/A	9.328e+00	7.98e-002 ug/g	
5	DUP	S98T002713	0 A QMSPR-A1 TC99-A1	SOLID	1.45e+00	1.58e+00	8.581 RPD	
5	DUP	S98T002713	0 A QMSPR-A1 SN117-A1	SOLID	<5.98e0	<5.87e0	RPD	
5	DUP	S98T002713	0 A QMSPR-A1 SN126-A1	SOLID	<5.98e0	<5.87e0	RPD	
5	DUP	S98T002713	0 A QMSPR-A1 CS133-A1	SOLID	3.04e+00	3.35e+00	9.703 RPD	
5	DUP	S98T002713	0 A QMSPR-A1 CS135-A1	SOLID	4.48e+00	5.21e+00	15.067 RPD	
5	DUP	S98T002713	0 A QMSPR-A1 CS137-A1	SOLID	9.33e+00	1.07e+01	13.679 RPD	
6	SAMPLE	S98T002846	0 A QMSPR-A1 TC99-A1	SOLID	N/A	1.701e+00	8.20e-002 ug/g	
6	SAMPLE	S98T002846	0 A QMSPR-A1 SN117-A1	SOLID	N/A	6.148e+00	6.149 ug/g	
6	SAMPLE	S98T002846	0 A QMSPR-A1 SN126-A1	SOLID	N/A	6.148e+00	6.149 ug/g	
6	SAMPLE	S98T002846	0 A QMSPR-A1 CS133-A1	SOLID	N/A	2.303e+00	8.20e-002 ug/g	
6	SAMPLE	S98T002846	0 A QMSPR-A1 CS135-A1	SOLID	N/A	2.574e+00	8.20e-002 ug/g	
6	SAMPLE	S98T002846	0 A QMSPR-A1 CS137-A1	SOLID	N/A	4.595e+00	8.20e-002 ug/g	
7	DUP	S98T002846	0 A QMSPR-A1 TC99-A1	SOLID	1.70e+00	1.65e+00	2.985 RPD	
7	DUP	S98T002846	0 A QMSPR-A1 SN117-A1	SOLID	<6.15e0	<6.09e0	RPD	
7	DUP	S98T002846	0 A QMSPR-A1 SN126-A1	SOLID	<6.15e0	<6.09e0	RPD	
7	DUP	S98T002846	0 A QMSPR-A1 CS133-A1	SOLID	2.30e+00	2.26e+00	1.754 RPD	
7	DUP	S98T002846	0 A QMSPR-A1 CS135-A1	SOLID	2.57e+00	2.59e+00	0.775 RPD	
7	DUP	S98T002846	0 A QMSPR-A1 CS137-A1	SOLID	4.59e+00	4.74e+00	3.215 RPD	
8	CCV	0	QMSPR-QC TC99	QC	0.02050	1.90e-02	92.683 % Recovery	
8	CCV	0	QMSPR-QC SN117	QC	0.018560	2.19e-02	117.996 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27746

Analyst: BTG Instrument: ICPMS1 Book# 48857, 49857Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSPR-QC	QC		
2 ICB			@MSPR-QC	QC		
3 BLNK-PREP			@MSPR-A1	SOLID		
4 SAMPLE	S98T002713 0 A		@MSPR-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
5 DUP	S98T002713 0 A		@MSPR-A1	SOLID		
6 SAMPLE	S98T002846 0 A		@MSPR-A1	SOLID	98000437	BY-105 (2)
	Analytes Requested: CS133-A1, CS135-A1, CS137-A1, SN117-A1, SN126-A1, TC99-A1					
7 DUP	S98T002846 0 A		@MSPR-A1	SOLID		
8 CCV			@MSPR-QC	QC		
9 CCB			@MSPR-QC	QC		

Final page for worklist # 27746

Signature

Date

Signature

Date

BTG
PREPBLK 1st-9th
S98T002713 .05-1-10-1
2713-D .05-1-10-1
2846 .05-1-10-1
2846-D .05-1-10-1

DF.

10

201

201

201

201

Data Entry Comments:

SCANS = PREPBLK - pb 2713, 2713, 2846

Upgraded & Validated by:

Saul M. Pung

03/10/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3018

HNF-1663 REV. 0

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/09/99 12:53 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	194000.
Stddev	1313.
%RSD	.6770

#1	195200.
#2	193700.
#3	193800.
#4	192900.
#5	192400.
#6	194700.
#7	192600.
#8	192900.
#9	195700.
#10	196000.

Bjg
3-9-99
BY-105
PRV.

588700 3384
3389
3389
2713
2846
2607
2608
2608

FILE# PR0308A.TXT

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3019 TO 3036

Method: PRV

Sample Name: Blank

Operator:

Comment:

Run Time: 03/09/99 12:55 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.645	54.70	1.504	25.00	1.851
Stddev	.198	3.77	.253	1.12	.222
%RSD	12.01	6.889	16.80	4.490	11.97

#1	1.617	54.00	1.516	25.58	1.895
#2	1.472	56.00	1.697	26.35	1.797
#3	1.830	57.00	1.103	24.25	2.181
#4	1.440	48.50	1.465	23.52	1.566
#5	1.867	58.00	1.741	25.33	1.816

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.123	2.566	3.144	10.64
Stddev	.313	.260	.217	.43
%RSD	14.72	10.11	6.900	4.023

#1	2.552	2.401	2.932	10.69
#2	1.947	2.296	3.394	10.11
#3	2.357	2.733	3.285	11.23
#4	1.920	2.930	2.905	10.79
#5	1.842	2.472	3.204	10.37

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.875
Stddev	.109
%RSD	.55001

#1	19.785
#2	20.035
#3	19.942
#4	19.795
#5	19.819

Method: PRV

Sample Name: PRV CAL

Operator:

Comment:

Run Time: 03/09/99

12:58 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Tc[99]	Cs[133]
Line	99/pulse	133/pulse
Units	Cts/S	Cts/S
Avg	21290.	18300.
Stddev	195.	113.
%RSD	.9168	.6163

#1	21020.	18180.
#2	21170.	18400.
#3	21340.	18400.
#4	21500.	18190.
#5	21430.	18340.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.733
Stddev	.073
%RSD	.37101

#1	19.771
#2	19.731
#3	19.618
#4	19.815
#5	19.729

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV1 CAL

Operator:

Comment:

Run Time: 03/09/99 13:00 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	Ru[102]	Te[125]	Cs[135]	Cs[137]
Line	102/pulse	125/pulse	135/pulse	137/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	703400.	1472.	6246.	10820.
Stddev	5886.	14.	59.	193.
%RSD	.8369	.9538	.9389	1.782

#1	710900.	1477.	6343.	11130.
#2	708400.	1493.	6241.	10650.
#3	700900.	1461.	6195.	10730.
#4	697800.	1470.	6248.	10860.
#5	698900.	1458.	6204.	10720.

Elem	Ba[138]
Line	138/pulse
Units	Cts/S
Avg	65850.
Stddev	1009.
%RSD	1.532

#1	67460.
#2	65280.
#3	65370.
#4	66180.
#5	64940.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.611
Stddev	.246
%RSD	1.2565

#1	19.254
#2	19.681
#3	19.590
#4	19.590
#5	19.942

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1080.8412	1.6454	1.0000000	03/09/99 12:59:27
Ru[102]	7033.2065	54.7000	1.0000000	03/09/99 13:02:01
Te[125]	14.7044	1.5045	1.0000000	03/09/99 13:02:01
Sn[117]	946.6558	25.0039	1.0000000	03/09/99 12:57:14
Sn[126]	947.8586	1.8512	1.0000000	03/09/99 12:57:14
Cs[133]	915.0640	2.1235	1.0000000	03/09/99 12:59:27
Cs[135]	945.9816	2.5664	1.0000000	03/09/99 13:02:01
Cs[137]	965.4728	3.1438	1.0000000	03/09/99 13:02:01
Ba[138]	658.3510	10.6369	1.0000000	03/09/99 13:02:01

HNF-1663 REV. 0

Method: PRV

Sample Name: PRV-SN CAL

Operator:

Comment:

Run Time: 03/09/99 13:13 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	Sn[117]	Sn[126]
Line	117/pulse	126/pulse
Units	Cts/S	Cts/S
Avg	16500.	16500.
Stddev	428.	428.
%RSD	2.594	2.594

#1	17150.	17150.
#2	16150.	16150.
#3	16510.	16510.
#4	16080.	16080.
#5	16640.	16640.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.257
Stddev	.334
%RSD	1.7336

#1	18.691
#2	19.388
#3	19.260
#4	19.387
#5	19.559

4NF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
Tc[99]	1080.8412	1.6454	1.0000000	03/09/99 12:59:27
Ru[102]	7033.2065	54.7000	1.0000000	03/09/99 13:02:01
Te[125]	14.7044	1.5045	1.0000000	03/09/99 13:02:01
Sn[117]	856.0545	25.0039	1.0000000	03/09/99 13:14:41
Sn[126]	857.2572	1.8512	1.0000000	03/09/99 13:14:41
Cs[133]	915.0640	2.1235	1.0000000	03/09/99 12:59:27
Cs[135]	945.9816	2.5664	1.0000000	03/09/99 13:02:01
Cs[137]	965.4728	3.1438	1.0000000	03/09/99 13:02:01
Ba[138]	658.3510	10.6369	1.0000000	03/09/99 13:02:01

Method: PRV Sample Name: icv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 13:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.23	.0005	.0124	1.418	.0003
Stddev	.49	.0011	.0195	.173	.0002
%RSD	2.564	236.9	157.1	12.22	71.03

#1	20.11	-.0007	-.0065	1.187	.0004
#2	19.14	.0003	.0161	1.302	.0001
#3	19.00	.0020	.0092	1.457	.0005
#4	19.01	-.0004	.0436	1.530	.0001
#5	18.92	.0012	-.0003	1.616	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.83	.0069	.0123	.1147
Stddev	.42	.0006	.0009	.0032
%RSD	2.105	8.250	7.566	2.802

#1	20.32	.0078	.0131	.1181
#2	19.17	.0065	.0115	.1153
#3	19.86	.0071	.0114	.1130
#4	19.97	.0064	.0119	.1101
#5	19.81	.0069	.0134	.1169

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.030
Stddev	.191
%RSD	1.0034

#1	18.819
#2	19.245
#3	19.205
#4	18.876
#5	19.005

Method: PRV Sample Name: icv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 13:25 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.1524	.0006	.0301	21.39	.0004
Stddev	.1494	.0010	.0091	.44	.0004
%RSD	98.04	160.1	30.32	2.044	104.6

#1	.4031	.0014	.0344	20.68	.0004
#2	.1731	-.0003	.0275	21.59	.0009
#3	.0919	-.0005	.0298	21.85	.0006
#4	.0563	.0009	.0419	21.42	.0003
#5	.0377	.0016	.0171	21.39	-.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0127	.0282	.0444	.4179
Stddev	.0031	.0018	.0013	.0213
%RSD	24.62	6.545	2.891	5.092

#1	.0177	.0267	.0455	.4072
#2	.0129	.0275	.0426	.4324
#3	.0123	.0301	.0453	.4473
#4	.0110	.0304	.0451	.3951
#5	.0095	.0266	.0434	.4075

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.733
Stddev	.596
%RSD	3.1788

#1	19.643
#2	18.864
#3	18.017
#4	18.630
#5	18.511

HNF-1663 REV. 0

Method: PRV

Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0010	.0063	-.0002	.0006
Stddev	.0002	.0010	.0265	.0007	.0005
%RSD	56.92	103.6	419.8	402.9	81.29

#1	.0003	.0010	-.0039	.0008	.0011
#2	.0001	-.0002	.0024	-.0005	-.0002
#3	.0005	.0020	.0213	.0002	.0005
#4	.0001	.0001	-.0292	-.0010	.0008
#5	.0004	.0019	.0411	-.0004	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0007	-.0002	-.0002	-.0002
Stddev	.0003	.0002	.0003	.0009
%RSD	41.85	68.20	173.5	533.4

#1	.0004	-.0002	-.0003	-.0005
#2	.0006	.0000	.0000	.0003
#3	.0008	-.0004	.0002	-.0002
#4	.0006	-.0003	-.0004	.0009
#5	.0012	-.0001	-.0003	-.0015

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.975
Stddev	.092
%RSD	.48244

#1	19.011
#2	18.827
#3	18.987
#4	19.076
#5	18.972

Method: PRV

Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/09/99 14:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0491	-.0003	.0119	.7644	.0000
Stddev	.0229	.0008	.0128	.0098	.0002
%RSD	46.70	268.9	107.2	1.288	577.7

#1	.0652	.0001	.0171	.7781	.0000
#2	.0773	-.0007	.0049	.7686	-.0001
#3	.0496	-.0016	.0319	.7653	-.0001
#4	.0319	.0003	.0050	.7535	.0000
#5	.0216	.0004	.0005	.7565	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0013	.0065	.0112	.1060
Stddev	.0003	.0005	.0004	.0042
%RSD	22.16	8.000	3.470	3.966

#1	.0013	.0073	.0112	.1127
#2	.0009	.0059	.0107	.1061
#3	.0017	.0067	.0110	.1019
#4	.0013	.0062	.0114	.1062
#5	.0012	.0063	.0118	.1030

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.787
Stddev	.424
%RSD	2.1424

#1	19.073
#2	19.978
#3	19.763
#4	19.954
#5	20.166

Method: PRV

Sample Name: s98t002713

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0364	.0336	.0013	.1273	.0003
Stddev	.0003	.0023	.0111	.0088	.0003
%RSD	.9406	6.754	855.6	6.902	100.7
#1	.0370	.0328	-.0099	.1423	.0001
#2	.0363	.0316	-.0039	.1248	.0009
#3	.0364	.0318	.0108	.1274	.0005
#4	.0361	.0351	.0154	.1200	.0001
#5	.0362	.0369	-.0058	.1222	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0761	.1123	.2339	1.339
Stddev	.0023	.0048	.0042	.024
%RSD	3.036	4.258	1.798	1.815
#1	.0800	.1183	.2402	1.375
#2	.0740	.1057	.2288	1.316
#3	.0760	.1105	.2325	1.333
#4	.0757	.1151	.2327	1.320
#5	.0749	.1120	.2353	1.353

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	20.144
Stddev	.193
%RSD	.95927

#1	19.870
#2	20.396
#3	20.144
#4	20.229
#5	20.084

Method: PRV

Sample Name: s98t002713_d

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0405	.0358	-.0045	.0302	.0007
Stddev	.0015	.0017	.0049	.0029	.0003
%RSD	3.775	4.750	107.7	9.461	42.74

#1	.0409	.0379	.0029	.0336	.0010
#2	.0421	.0332	-.0107	.0292	.0006
#3	.0399	.0362	-.0050	.0324	.0008
#4	.0381	.0352	-.0055	.0264	.0004
#5	.0413	.0363	-.0042	.0294	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0857	.1332	.2745	1.657
Stddev	.0022	.0016	.0049	.034
%RSD	2.549	1.191	1.787	2.058

#1	.0844	.1312	.2704	1.630
#2	.0833	.1347	.2725	1.698
#3	.0891	.1341	.2812	1.684
#4	.0855	.1343	.2702	1.618
#5	.0861	.1319	.2780	1.653

Int. Std.	Int[115]
Line	115/pulse
Units	Cts/S
Avg	19.391
Stddev	.328
%RSD	1.6940

#1	19.713
#2	18.928
#3	19.216
#4	19.678
#5	19.418

Method: PRV

Sample Name: s98t002846

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0415	.0506	.0061	.0284	.0005
Stddev	.0016	.0022	.0120	.0022	.0002
%RSD	3.897	4.317	195.5	7.669	44.72

#1	.0432	.0492	.0139	.0299	.0007
#2	.0408	.0525	.0120	.0288	.0004
#3	.0395	.0501	.0029	.0250	.0001
#4	.0431	.0480	.0152	.0279	.0007
#5	.0408	.0531	-.0135	.0307	.0006

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0562	.0628	.1121	.6908
Stddev	.0011	.0020	.0023	.0108
%RSD	2.004	3.153	2.019	1.564

#1	.0551	.0623	.1095	.7030
#2	.0575	.0649	.1118	.6744
#3	.0552	.0598	.1118	.6935
#4	.0558	.0628	.1119	.6871
#5	.0572	.0642	.1158	.6962

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.493
Stddev	.257
%RSD	1.3192

#1	19.308
#2	19.625
#3	19.714
#4	19.680
#5	19.134

Method: PRV

Sample Name: s98t002846_d

Operator: bjg

Comment: 201 df

Run Time: 03/09/99 14:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0407	.0512	-.0005	.0263	.0004
Stddev	.0015	.0018	.0144	.0022	.0002
%RSD	3.595	3.446	2823.	8.367	56.39

#1	.0403	.0517	-.0029	.0287	.0004
#2	.0416	.0481	.0028	.0287	.0007
#3	.0396	.0521	-.0132	.0252	.0004
#4	.0429	.0524	.0225	.0239	.0002
#5	.0394	.0519	-.0117	.0251	.0002

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0557	.0638	.1168	.7346
Stddev	.0031	.0018	.0026	.0147
%RSD	5.488	2.785	2.203	1.998

#1	.0600	.0644	.1204	.7515
#2	.0563	.0642	.1163	.7299
#3	.0547	.0610	.1132	.7121
#4	.0515	.0635	.1167	.7400
#5	.0557	.0658	.1175	.7393

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.674
Stddev	.374
%RSD	1.8996

#1	19.502
#2	19.409
#3	20.224
#4	19.344
#5	19.892

Method: PRV Sample Name: ccv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 15:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0016	-.0004	.0198	21.86	.0000
Stddev	.0007	.0009	.0098	1.15	.000
%RSD	42.70	246.0	49.38	5.273	3962.

#1	.0024	-.0010	.0188	22.36	.0000
#2	.0015	-.0012	.0225	21.44	-.0004
#3	.0020	-.0007	.0347	21.70	-.0003
#4	.0014	.0009	.0092	23.46	.0007
#5	.0006	.0003	.0137	20.34	.0000

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0073	.0282	.0447	.4165
Stddev	.0012	.0013	.0024	.0073
%RSD	16.55	4.578	5.355	1.752

#1	.0079	.0275	.0463	.4204
#2	.0065	.0271	.0430	.4113
#3	.0059	.0290	.0445	.4191
#4	.0090	.0301	.0477	.4248
#5	.0074	.0273	.0418	.4067

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.205
Stddev	.591
%RSD	3.2485

#1	17.961
#2	18.800
#3	18.115
#4	17.388
#5	18.762

Method: PRV Sample Name: ccv

Operator: bjg

Comment: msprv lcs

Run Time: 03/09/99 15:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	19.03	.0032	.0035	1.443	.0001
Stddev	.88	.0003	.0127	.158	.0002
%RSD	4.640	9.593	365.7	10.94	330.8

#1	17.46	.0036	.0004	1.195	-.0003
#2	19.40	.0030	-.0084	1.392	.0002
#3	19.58	.0035	.0250	1.486	.0002
#4	19.41	.0032	.0028	1.544	.0000
#5	19.32	.0029	-.0024	1.596	.0001

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	19.53	.0069	.0121	.1130
Stddev	.74	.0006	.0008	.0044
%RSD	3.766	9.131	6.314	3.885

#1	18.25	.0060	.0112	.1053
#2	20.02	.0072	.0130	.1155
#3	19.89	.0076	.0122	.1147
#4	19.55	.0073	.0114	.1134
#5	19.93	.0066	.0126	.1160

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	19.549
Stddev	.746
%RSD	3.8142

#1	20.863
#2	19.191
#3	19.233
#4	19.408
#5	19.049

Method: PRV

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 15:36 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	Tc[99]	Ru[102]	Te[125]	Sn[117]	Sn[126]
Line	99/pulse	102/pulse	125/pulse	117/pulse	126/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0015	.0010	-.0009	-.0009	.0001
Stddev	.0001	.0015	.0064	.0016	.0004
%RSD	4.753	157.1	721.0	177.1	368.0

#1	.0014	-.0001	.0061	-.0024	.0001
#2	.0015	-.0007	-.0018	-.0012	.0002
#3	.0016	.0017	.0047	.0012	-.0005
#4	.0015	.0031	-.0041	-.0025	.0003
#5	.0015	.0008	-.0093	.0004	.0004

Elem	Cs[133]	Cs[135]	Cs[137]	Ba[138]
Line	133/pulse	135/pulse	137/pulse	138/pulse
Units	ppb	ppb	ppb	ppb
Avg	.0006	-.0001	-.0001	-.0005
Stddev	.0005	.0002	.0003	.0012
%RSD	72.24	164.3	364.9	271.8

#1	.0002	.0001	.0001	.0011
#2	.0013	-.0001	-.0002	.0002
#3	.0009	.0000	-.0005	-.0016
#4	.0006	-.0002	-.0001	.0000
#5	.0002	-.0004	.0003	-.0018

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	18.583
Stddev	.289
%RSD	1.5544

#1	18.184
#2	18.948
#3	18.431
#4	18.698
#5	18.653

LABCORE Completed Worklist Report for Worklist# 27946

Analyst: bjg

Instrument: ICPMS1

Book#: 9859

Method: LA-506-101 Rev/Mod A-Ø

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	ØMSAC-QC TH232	QC	1.00e-3	1.08e-02	108.000 % Recovery	
1	ICV	0	ØMSAC-QC NP237	QC	9.84e-3	1.14e-02	115.854 % Recovery	
1	ICV	0	ØMSAC-QC PU239	QC	1.00e-2	1.01e-02	101.000 % Recovery	
1	ICV	0	ØMSAC-QC 241	QC	1.03e-2	1.07e-02	103.883 % Recovery	
2	ICB	0	ØMSAC-QC PA231	QC	1	<3.21e-6		ug/mL
2	ICB	0	ØMSAC-QC TH229	QC	1	<3.19e-6		ug/mL
2	ICB	0	ØMSAC-QC TH230	QC	1	<3.00e-6		ug/mL
2	ICB	0	ØMSAC-QC TH232	QC	1	<4.62e-6		ug/mL
2	ICB	0	ØMSAC-QC NP237	QC	1	<3.14e-6		ug/mL
2	ICB	0	ØMSAC-QC PU239	QC	1	<2.82e-6		ug/mL
2	ICB	0	ØMSAC-QC PU240	QC	1	<4.27e-6		ug/mL
2	ICB	0	ØMSAC-QC PU242	QC	1	<3.31e-6		ug/mL
2	ICB	0	ØMSAC-QC 241	QC	1	<3.26e-6		ug/mL
2	ICB	0	ØMSAC-QC 243	QC	1	<4.06e-6		ug/mL
2	ICB	0	ØMSAC-QC 244	QC	1	<5.88e-6		ug/mL
3	BLNK-PREP	0	ØMSACTAL PA231-A1	SOLID	1	<3.53e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL TH229-A1	SOLID	1	<3.51e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL TH230-A1	SOLID	1	<3.29e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL TH232-A1	SOLID	1	<5.08e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL NP237-A1	SOLID	1	<3.45e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL PU239-A1	SOLID	1	<3.10e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL PU240-A1	SOLID	1	<4.70e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL PU242-A1	SOLID	1	<3.64e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL 241-A1	SOLID	1	<3.59e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL 243-A1	SOLID	1	<4.47e-5		ug/g
3	BLNK-PREP	0	ØMSACTAL 244-A1	SOLID	1	<6.47e-5		ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL PA231-A1	SOLID	N/A	< 1.188e-01	0.119	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL TH229-A1	SOLID	N/A	< 1.183e-01	0.118	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL TH230-A1	SOLID	N/A	< 1.109e-01	0.111	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL TH232-A1	SOLID	N/A	< 2.131e-01	0.171	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL NP237-A1	SOLID	N/A	< 1.162e-01	0.116	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL PU239-A1	SOLID	N/A	< 1.043e-01	0.104	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL PU240-A1	SOLID	N/A	< 1.581e-01	0.158	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL PU242-A1	SOLID	N/A	< 1.225e-01	0.122	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL 241-A1	SOLID	N/A	< 1.207e-01	0.121	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL 243-A1	SOLID	N/A	< 1.505e-01	0.150	ug/g
4	SAMPLE	S98T002604	0 A ØMSACTAL 244-A1	SOLID	N/A	< 2.179e-01	0.218	ug/g
5	DUP	S98T002604	0 A ØMSACTAL PA231-A1	SOLID	<1.19e-1	<1.15e-1		RPD
5	DUP	S98T002604	0 A ØMSACTAL TH229-A1	SOLID	<1.18e-1	<1.15e-1		RPD
5	DUP	S98T002604	0 A ØMSACTAL TH230-A1	SOLID	<1.11e-1	<1.08e-1		RPD
5	DUP	S98T002604	0 A ØMSACTAL TH232-A1	SOLID	2.13e-01	<1.66e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27946

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 DUP	S98T002604	0 A	OMSACT1 NP237-A1	SOLID	<1.16e-1	<1.13e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 PU239-A1	SOLID	<1.04e-1	<1.01e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 PU240-A1	SOLID	<1.58e-1	<1.53e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 PU242-A1	SOLID	<1.22e-1	<1.19e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 241-A1	SOLID	<1.21e-1	<1.17e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 243-A1	SOLID	<1.50e-1	<1.46e-1		RPD
5 DUP	S98T002604	0 A	OMSACT1 244-A1	SOLID	<2.18e-1	<2.11e-1		RPD
6 SPK-POST	S98T002604	0 A	OMSACT1 TH232-A1	SOLID	0.020	2.03e-02	101.500 % Recovery	
6 SPK-POST	S98T002604	0 A	OMSACT1 NP237-A1	SOLID	0.01976	2.25e-02	113.866 % Recovery	
6 SPK-POST	S98T002604	0 A	OMSACT1 PU239-A1	SOLID	0.02008	1.97e-02	98.108 % Recovery	
6 SPK-POST	S98T002604	0 A	OMSACT1 241-A1	SOLID	0.020520	2.12e-02	103.314 % Recovery	
7 SAMPLE	S98T002605	0 A	OMSACT1 PA231-A1	SOLID	N/A	< 1.145e-01	0.115	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 TH229-A1	SOLID	N/A	< 1.140e-01	0.114	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 TH230-A1	SOLID	N/A	< 1.069e-01	0.107	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 TH232-A1	SOLID	N/A	< 1.649e-01	0.165	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 NP237-A1	SOLID	N/A	< 1.120e-01	0.112	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 PU239-A1	SOLID	N/A	< 1.005e-01	0.101	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 PU240-A1	SOLID	N/A	< 1.524e-01	0.152	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 PU242-A1	SOLID	N/A	< 1.181e-01	0.118	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 241-A1	SOLID	N/A	< 1.163e-01	0.116	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 243-A1	SOLID	N/A	< 1.451e-01	0.145	ug/g
7 SAMPLE	S98T002605	0 A	OMSACT1 244-A1	SOLID	N/A	< 2.100e-01	0.210	ug/g
8 DUP	S98T002605	0 A	OMSACT1 PA231-A1	SOLID	<1.15e-1	<1.18e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 TH229-A1	SOLID	<1.14e-1	<1.18e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 TH230-A1	SOLID	<1.07e-1	<1.10e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 TH232-A1	SOLID	<1.65e-1	<1.70e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 NP237-A1	SOLID	<1.12e-1	<1.16e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 PU239-A1	SOLID	<1.01e-1	<1.04e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 PU240-A1	SOLID	<1.52e-1	<1.57e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 PU242-A1	SOLID	<1.18e-1	<1.22e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 241-A1	SOLID	<1.16e-1	<1.20e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 243-A1	SOLID	<1.45e-1	<1.50e-1		RPD
8 DUP	S98T002605	0 A	OMSACT1 244-A1	SOLID	<2.10e-1	<2.17e-1		RPD
9 SAMPLE	S98T002606	0 A	OMSACT1 PA231-A1	SOLID	N/A	< 1.209e-01	0.121	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 TH229-A1	SOLID	N/A	< 1.204e-01	0.120	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 TH230-A1	SOLID	N/A	< 1.129e-01	0.113	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 TH232-A1	SOLID	N/A	6.068e+00	0.174	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 NP237-A1	SOLID	N/A	< 1.182e-01	0.118	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 PU239-A1	SOLID	N/A	< 1.062e-01	0.106	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 PU240-A1	SOLID	N/A	< 1.609e-01	0.161	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 PU242-A1	SOLID	N/A	< 1.247e-01	0.125	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 241-A1	SOLID	N/A	< 1.228e-01	0.123	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 243-A1	SOLID	N/A	< 1.532e-01	0.153	ug/g
9 SAMPLE	S98T002606	0 A	OMSACT1 244-A1	SOLID	N/A	< 2.217e-01	0.222	ug/g
10 DUP	S98T002606	0 A	OMSACT1 PA231-A1	SOLID	<1.21e-1	<1.19e-1		RPD
10 DUP	S98T002606	0 A	OMSACT1 TH229-A1	SOLID	<1.20e-1	<1.19e-1		RPD
10 DUP	S98T002606	0 A	OMSACT1 TH230-A1	SOLID	<1.13e-1	<1.11e-1		RPD
10 DUP	S98T002606	0 A	OMSACT1 TH232-A1	SOLID	6.07e+00	7.12e+00	15.921	RPD
10 DUP	S98T002606	0 A	OMSACT1 NP237-A1	SOLID	<1.18e-1	<1.16e-1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27946

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
10 DUP	S98T002606	0 A	QMSACTA1 FU239-A1	SOLID	<1.06e-1	<1.05e-1		RPD
10 DUP	S98T002606	0 A	QMSACTA1 FU240-A1	SOLID	<1.61e-1	<1.58e-1		RPD
10 DUP	S98T002606	0 A	QMSACTA1 FU242-A1	SOLID	<1.25e-1	<1.23e-1		RPD
10 DUP	S98T002606	0 A	QMSACTA1 241-A1	SOLID	<1.23e-1	<1.21e-1		RPD
10 DUP	S98T002606	0 A	QMSACTA1 243-A1	SOLID	<1.53e-1	<1.51e-1		RPD
10 DUP	S98T002606	0 A	QMSACTA1 244-A1	SOLID	<2.22e-1	<2.18e-1		RPD
11 CCV	0	0	QMSAC-QC TH232	QC	1.0e-2	1.06e-02	106.000 % Recovery	
11 CCV	0	0	QMSAC-QC NP237	QC	9.84e-3	1.19e-02	120.935 % Recovery	
11 CCV	0	0	QMSAC-QC FU239	QC	1.00e-2	1.05e-02	105.000 % Recovery	
11 CCV	0	0	QMSAC-QC 241	QC	1.03e-2	1.12e-02	108.738 % Recovery	
12 CCB	0	0	QMSAC-QC FA231	QC	1	<3.21e-6		ug/mL
12 CCB	0	0	QMSAC-QC TH229	QC	1	<3.19e-6		ug/mL
12 CCB	0	0	QMSAC-QC TH230	QC	1	<3.00e-6		ug/mL
12 CCB	0	0	QMSAC-QC TH232	QC	1	<4.62e-6		ug/mL
12 CCB	0	0	QMSAC-QC NP237	QC	1	<3.14e-6		ug/mL
12 CCB	0	0	QMSAC-QC FU239	QC	1	<2.82e-6		ug/mL
12 CCB	0	0	QMSAC-QC FU240	QC	1	<4.27e-6		ug/mL
12 CCB	0	0	QMSAC-QC FU242	QC	1	<3.31e-6		ug/mL
12 CCB	0	0	QMSAC-QC 241	QC	1	<3.26e-6		ug/mL
12 CCB	0	0	QMSAC-QC 243	QC	1	<4.06e-6		ug/mL
12 CCB	0	0	QMSAC-QC 244	QC	1	<5.88e-6		ug/mL

Final page for worklist# 27946

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature Date 3-3-99

LABCORE Data Entry Template for Worklist# 27946

Analyst: KST Instrument: ICPMS1 Book# 9855

Method: LA-506-101 Rev/Mod A-0 9859

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSAC-QC	QC		
2 ICB			@MSAC-QC	QC		
3 BLNK-PREP			@MSACTA1	SOLID		
4 SAMPLE	S98T002604	0 A	@MSACTA1	SOLID	98000436	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
5 DUP	S98T002604	0 A	@MSACTA1	SOLID		
6 SPK-POST	S98T002604	0 A	@MSACTA1	SOLID		
7 SAMPLE	S98T002605	0 A	@MSACTA1	SOLID	98000436	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
8 DUP	S98T002605	0 A	@MSACTA1	SOLID		
9 SAMPLE	S98T002606	0 A	@MSACTA1	SOLID	98000436	BY-105 (2)
	Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1					
10 DUP	S98T002606	0 A	@MSACTA1	SOLID		
11 CCV			@MSAC-QC	QC		
12 CCB			@MSAC-QC	QC		

Final page for worklist # 27946

KS Romoike 02-28-99
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: Sample-0

Operator:

Comment:

Run Time: 02/28/99 10:31 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	171100.
Stddev	1597.
%RSD	.9335

#1	169700.
#2	171100.
#3	169800.
#4	170900.
#5	169400.
#6	169700.
#7	171700.
#8	173400.
#9	171400.
#10	174100.

SP
2/28/99

File ac0228a.txt

Blnk- prep 1-10 11
S0987002604 .05-10 001
2604-d
2604-a
2605-
2605-d
2606-
2606-d

04 ml of 8BSS

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3041 TO 3069

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.010 ppm cal

Run Time: 02/28/99 10:59 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2408	.3076	448.8	469.1	.8543
Stddev	.0189	.0261	3.9	7.2	.0541
%RSD	7.856	8.478	.8611	1.541	6.331

#1	.2419	.3393	447.5	474.3	.9449
#2	.2712	.3279	448.0	473.1	.8135
#3	.2292	.3065	443.9	473.1	.8612
#4	.2404	.2811	449.9	456.9	.8355
#5	.2215	.2831	454.5	467.9	.8165

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.005	.3790	.2568	.2780	.2639
Stddev	.057	.0427	.0333	.0168	.0265
%RSD	5.658	11.27	12.96	6.026	10.05

#1	1.002	.4066	.2405	.3006	.3092
#2	.9366	.4344	.3002	.2739	.2449
#3	1.025	.3784	.2197	.2645	.2509
#4	.9734	.3442	.2824	.2614	.2654
#5	1.088	.3316	.2411	.2896	.2490

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2633	.2549
Stddev	.0183	.0333
%RSD	6.934	13.06

#1	.2849	.3064
#2	.2490	.2310
#3	.2685	.2699
#4	.2404	.2273
#5	.2739	.2398

Sample Name: cal std Run Time: 02/28/99 10:59

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	245.53
Stddev	9.05
%RSD	3.6860

#1	232.83
#2	240.93
#3	245.79
#4	253.74
#5	254.35

Method: ACTINIDE Sample Name: cal std

Operator: kjt

Comment: 0.100 ppm

Run Time: 02/28/99 11:07 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2873	.6892	.4399.	.4484.	6.501
Stddev	.0224	.0532	81.	106.	.086
%RSD	7.790	7.722	1.848	2.375	1.318

#1	.2637	.6135	4458.	4545.	6.493
#2	.2955	.6522	4375.	4464.	6.543
#3	.3208	.7263	4272.	4381.	6.357
#4	.2849	.7276	4415.	4394.	6.536
#5	.2714	.7264	4476.	4634.	6.576

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7.603	1.698	.2486	.2536	.2630
Stddev	.115	.061	.0180	.0181	.0191
%RSD	1.513	3.579	7.242	7.139	7.265

#1	7.731	1.746	.2624	.2435	.2812
#2	7.461	1.597	.2630	.2630	.2812
#3	7.504	1.708	.2335	.2759	.2400
#4	7.664	1.695	.2592	.2566	.2656
#5	7.657	1.745	.2249	.2288	.2469

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2468	.2385
Stddev	.0285	.0154
%RSD	11.53	6.471

#1	.2597	.2583
#2	.2617	.2500
#3	.2631	.2207
#4	.1963	.2284
#5	.2534	.2353

Sample Name: cal std Run Time: 02/28/99 11:07

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	256.24
Stddev	4.99
%RSD	1.9479

#1	247.75
#2	256.04
#3	259.77
#4	259.76
#5	257.88

Method: ACTINIDE Sample Name: icv

Operator: kjt

Comment: lcs icv std

Run Time: 02/28/99 11:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2490	.3034	484.2	19.78	.9376
Stddev	.0164	.0199	11.3	.34	.0384
%RSD	6.574	6.551	2.338	1.738	4.094

#1	.2724	.2966	471.9	19.29	.8831
#2	.2354	.2890	486.2	20.22	.9276
#3	.2358	.2830	474.6	19.67	.9460
#4	.2418	.3192	500.0	19.92	.9411
#5	.2597	.3291	488.2	19.80	.9900

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	532.8	478.1	12.89	.3469	.2695
Stddev	11.3	7.8	.33	.0249	.0128
%RSD	2.123	1.634	2.551	7.183	4.764

#1	517.3	471.1	12.93	.3195	.2831
#2	548.5	488.8	12.94	.3331	.2698
#3	529.7	474.8	12.41	.3840	.2695
#4	536.8	483.7	13.34	.3400	.2764
#5	531.7	471.9	12.85	.3577	.2489

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	514.9	.2654
Stddev	4.3	.0199
%RSD	.8289	7.500

#1	515.2	.2656
#2	513.8	.2381
#3	512.5	.2897
#4	522.0	.2778
#5	511.0	.2557

Sample Name: icv Run Time: 02/28/99 11:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	244.62
Stddev	2.83
%RSD	1.1589

#1	247.23
#2	242.19
#3	247.37
#4	241.20
#5	245.12

Method: ACTINIDE Sample Name: icb

Operator: kjt

Comment: 2% hno3

Run Time: 02/28/99 11:23 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2473	.2471	.4274	.2878	.2391
Stddev	.0055	.0125	.0388	.0265	.0290
%RSD	2.224	5.053	9.081	9.193	12.13

#1	.2456	.2509	.4444	.2763	.2389
#2	.2508	.2604	.4811	.2947	.2728
#3	.2508	.2280	.4131	.3125	.2092
#4	.2510	.2537	.4219	.3079	.2632
#5	.2384	.2424	.3763	.2477	.2116

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2532	.3658	.2618	.2416	.2679
Stddev	.0260	.0194	.0122	.0117	.0237
%RSD	10.27	5.300	4.674	4.839	8.857

#1	.2255	.3710	.2509	.2562	.2269
#2	.2344	.3399	.2577	.2508	.2796
#3	.2843	.3929	.2722	.2293	.2776
#4	.2767	.3676	.2510	.2320	.2862
#5	.2451	.3575	.2772	.2397	.2692

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2715	.2655
Stddev	.0066	.0239
%RSD	2.444	9.014

#1	.2789	.2389
#2	.2700	.2920
#3	.2615	.2494
#4	.2713	.2578
#5	.2759	.2892

Sample Name: icb Run Time: 02/28/99 11:23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	247.23
Stddev	2.73
%RSD	1.1032

#1	249.76
#2	243.18
#3	248.56
#4	245.71
#5	248.92

Method: ACTINIDE Sample Name: blnk-prep

Operator: kjt

Comment: 11 df

Run Time: 02/28/99 11:27 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2561	.2618	.4503	.3608	.2471
Stddev	.0292	.0133	.0310	.0238	.0110
%RSD	11.42	5.088	6.878	6.608	4.443
#1	.2776	.2762	.4943	.3951	.2521
#2	.2804	.2570	.4711	.3440	.2473
#3	.2566	.2635	.4359	.3379	.2345
#4	.2582	.2708	.4243	.3517	.2624
#5	.2075	.2419	.4260	.3752	.2391

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2785	.3758	.2580	.2737	.2687
Stddev	.0127	.0346	.0196	.0234	.0196
%RSD	4.566	9.203	7.605	8.549	7.284
#1	.2960	.4291	.2648	.2974	.2847
#2	.2735	.3633	.2611	.2846	.2487
#3	.2662	.3779	.2538	.2621	.2497
#4	.2875	.3336	.2820	.2387	.2694
#5	.2694	.3752	.2281	.2859	.2914

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2839	.2695
Stddev	.0196	.0240
%RSD	6.912	8.920
#1	.2677	.2578
#2	.3177	.2611
#3	.2745	.2414
#4	.2820	.2847
#5	.2776	.3024

Sample Name: blk-prep Run Time: 02/28/99 11:27

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	239.93
Stddev	2.91
%RSD	1.2113

#1	235.36
#2	241.29
#3	241.66
#4	238.81
#5	242.54

Method: ACTINIDE Sample Name: s98t002604

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 11:31 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2434	.2389	.5748	1.228	.2526
Stddev	.0138	.0174	.0291	.032	.0141
%RSD	5.686	7.283	5.060	2.566	5.596

#1	.2312	.2257	.5951	1.198	.2408
#2	.2557	.2640	.5770	1.280	.2501
#3	.2291	.2207	.5833	1.227	.2586
#4	.2416	.2472	.5942	1.225	.2739
#5	.2595	.2371	.5246	1.208	.2399

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2623	.3111	.2614	.2602	.2561
Stddev	.0262	.0174	.0177	.0094	.0280
%RSD	9.997	5.603	6.766	3.612	10.94

#1	.2709	.3324	.2804	.2449	.2257
#2	.2333	.3101	.2501	.2654	.2529
#3	.2994	.3078	.2811	.2600	.3022
#4	.2669	.2852	.2472	.2697	.2515
#5	.2413	.3198	.2483	.2609	.2483

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2532	.2584
Stddev	.0343	.0160
%RSD	13.54	6.183

#1	.2490	.2572
#2	.2193	.2626
#3	.2277	.2727
#4	.2641	.2318
#5	.3058	.2679

HNF-1663 REV. 0

Sample Name: s98t002604 Run Time: 02/28/99 11:31

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	238.87
Stddev	2.73
%RSD	1.1446

#1	243.66
#2	238.60
#3	237.18
#4	237.29
#5	237.64

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002604_d

Operator: Kjt

Comment: 201 df

Run Time: 02/28/99 11:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2394	.2300	.4968	1.595	.2505
Stddev	.0191	.0266	.0306	.042	.0180
%RSD	7.984	11.58	6.166	2.617	7.168
#1	.2309	.2169	.5066	1.538	.2449
#2	.2682	.2393	.4758	1.573	.2737
#3	.2171	.2644	.4829	1.593	.2296
#4	.2460	.1932	.5462	1.632	.2640
#5	.2349	.2363	.4726	1.639	.2404

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2559	.2911	.2617	.2641	.2586
Stddev	.0245	.0224	.0102	.0151	.0145
%RSD	9.563	7.710	3.904	5.711	5.591
#1	.2757	.3261	.2575	.2449	.2379
#2	.2475	.2929	.2613	.2847	.2627
#3	.2435	.2839	.2686	.2574	.2547
#4	.2863	.2640	.2738	.2613	.2779
#5	.2266	.2888	.2473	.2722	.2598

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2440	.2614
Stddev	.0202	.0098
%RSD	8.261	3.743
#1	.2407	.2729
#2	.2132	.2709
#3	.2616	.2519
#4	.2627	.2543
#5	.2418	.2570

Sample Name: s98t002604_d Run Time: 02/28/99 11:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	240.24
Stddev	1.61
%RSD	.67222

#1	238.19
#2	242.38
#3	239.53
#4	239.85
#5	241.24

Method: ACTINIDE Sample Name: s98t002604_a

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 11:40 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2671	.3452	909.0	3.098	1.442
Stddev	.0184	.0264	8.9	.066	.102
%RSD	6.902	7.647	.9787	2.137	7.086

#1	.2456	.3194	905.7	3.188	1.289
#2	.2522	.3784	901.9	3.037	1.502
#3	.2668	.3250	901.7	3.082	1.416
#4	.2876	.3679	913.0	3.039	1.560
#5	.2831	.3353	922.6	3.143	1.444

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1046.	929.0	25.29	.3819	.2793
Stddev	21.	21.0	.43	.0226	.0383
%RSD	1.980	2.261	1.709	5.927	13.72

#1	1044.	909.5	25.22	.3946	.2427
#2	1031.	914.3	24.71	.3415	.3373
#3	1024.	917.5	25.21	.3917	.2753
#4	1076.	952.1	25.39	.3898	.2482
#5	1057.	951.4	25.91	.3919	.2932

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1015.	.2868
Stddev	20.	.0155
%RSD	1.922	5.410

#1	1017.	.2995
#2	999.8	.2919
#3	990.4	.2895
#4	1037.	.2598
#5	1029.	.2932

Sample Name: s98t002604_a Run Time: 02/28/99 11:40

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	232.58
Stddev	3.31
%RSD	1.4253

#1	234.84
#2	235.22
#3	234.88
#4	228.35
#5	229.63

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002605

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 11:48 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2486	.2654	.5142	5.239	.2760
Stddev	.0273	.0091	.0361	.144	.0250
%RSD	10.97	3.423	7.028	2.753	9.069
#1	.2783	.2599	.5495	5.197	.2825
#2	.2359	.2630	.5560	5.403	.2716
#3	.2131	.2798	.4801	5.099	.2500
#4	.2421	.2564	.5014	5.379	.2607
#5	.2736	.2679	.4842	5.116	.3152

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2823	.6485	.2865	.2705	.2768
Stddev	.0258	.0243	.0225	.0187	.0188
%RSD	9.144	3.752	7.858	6.925	6.778
#1	.2543	.6851	.2769	.2995	.2853
#2	.3216	.6475	.2959	.2559	.2587
#3	.2741	.6548	.3210	.2599	.2599
#4	.2922	.6203	.2751	.2579	.2765
#5	.2693	.6346	.2636	.2793	.3037

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2924	.2586
Stddev	.0136	.0127
%RSD	4.642	4.926
#1	.3079	.2571
#2	.2887	.2730
#3	.3054	.2514
#4	.2822	.2421
#5	.2779	.2693

HNF-1663 REV. 0

Sample Name: s98t002605 Run Time: 02/28/99 11:48

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	233.84
Stddev	1.44
%RSD	.61765

#1	235.97
#2	233.20
#3	234.68
#4	232.68
#5	232.68

Method: ACTINIDE Sample Name: s98t002605_d

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 11:52 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2632	.2481	.4516	5.165	.2595
Stddev	.0201	.0242	.0351	.086	.0248
%RSD	7.646	9.766	7.770	1.669	9.556
#1	.2821	.2706	.4811	5.105	.2620
#2	.2288	.2573	.4179	5.121	.2899
#3	.2697	.2301	.4778	5.293	.2609
#4	.2671	.2671	.4091	5.092	.2207
#5	.2681	.2153	.4720	5.214	.2638

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2739	.5074	.2823	.2804	.2597
Stddev	.0384	.0369	.0095	.0256	.0131
%RSD	14.03	7.266	3.350	9.134	5.043
#1	.3150	.5570	.2835	.2692	.2420
#2	.2530	.4946	.2871	.2928	.2757
#3	.3093	.5218	.2814	.3151	.2594
#4	.2685	.5075	.2671	.2783	.2530
#5	.2239	.4564	.2924	.2467	.2681

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2869	.2680
Stddev	.0338	.0173
%RSD	11.78	6.470
#1	.2835	.2806
#2	.3440	.2558
#3	.2829	.2756
#4	.2572	.2840
#5	.2667	.2439

Sample Name: s98t002605_d Run Time: 02/28/99 11:52

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	233.12
Stddev	3.57
%RSD	1.5328

#1	232.79
#2	234.53
#3	227.41
#4	237.13
#5	233.74

Method: ACTINIDE Sample Name: s98t002606

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 11:56 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2570	.2478	7.538	346.1	.2571
Stddev	.0175	.0169	.032	5.0	.0130
%RSD	6.792	6.828	.4292	1.454	5.043

#1	.2274	.2389	7.575	344.8	.2590
#2	.2674	.2745	7.489	339.6	.2432
#3	.2583	.2312	7.545	345.1	.2469
#4	.2719	.2532	7.526	347.4	.2762
#5	.2600	.2410	7.554	353.5	.2600

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8625	.7005	.2901	.2522	.2498
Stddev	.0396	.0473	.0130	.0226	.0105
%RSD	4.586	6.751	4.472	8.954	4.212

#1	.8591	.7569	.2777	.2374	.2576
#2	.8435	.7326	.2959	.2333	.2617
#3	.8092	.7036	.2869	.2712	.2383
#4	.8904	.6387	.2805	.2373	.2517
#5	.9101	.6705	.3097	.2819	.2396

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2880	.2676
Stddev	.0247	.0293
%RSD	8.578	10.95

#1	.2777	.2820
#2	.2688	.2347
#3	.2712	.2369
#4	.2934	.2920
#5	.3287	.2922

Sample Name: s98t002606 Run Time: 02/28/99 11:56

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	231.89
Stddev	2.38
%RSD	1.0258

#1	231.64
#2	234.33
#3	233.57
#4	231.74
#5	228.18

Method: ACTINIDE Sample Name: s98t002606_d

Operator: kjt

Comment: 201 df

Run Time: 02/28/99 12:00 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2464	.2573	8.917	402.3	.2593
Stddev	.0260	.0250	.151	5.5	.0193
%RSD	10.55	9.729	1.689	1.361	7.450

#1	.2273	.2776	8.975	406.0	.2690
#2	.2599	.2585	9.143	409.4	.2542
#3	.2210	.2210	8.756	395.8	.2595
#4	.2391	.2829	8.892	398.7	.2829
#5	.2846	.2463	8.818	401.6	.2308

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.8884	.6971	.2786	.2529	.2645
Stddev	.0413	.0461	.0194	.0138	.0274
%RSD	4.651	6.606	6.963	5.460	10.36

#1	.8587	.7235	.2690	.2618	.2546
#2	.9502	.7221	.2773	.2383	.2253
#3	.8898	.6503	.2524	.2510	.2909
#4	.8997	.6451	.3013	.2716	.2617
#5	.8437	.7447	.2930	.2421	.2902

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2810	.2635
Stddev	.0114	.0322
%RSD	4.057	12.22

#1	.2762	.3049
#2	.2989	.2874
#3	.2852	.2381
#4	.2702	.2292
#5	.2746	.2577

Sample Name: s98t002606_d Run Time: 02/28/99 12:00

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	233.48
Stddev	2.16
%RSD	.92618

#1	231.75
#2	230.82
#3	233.76
#4	235.64
#5	235.46

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: ccv

Operator: kjt

Comment: lcs std

Run Time: 02/28/99 12:07 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2599	.2853	477.6	19.99	.8633
Stddev	.0337	.0165	7.8	.34	.0374
%RSD	12.99	5.782	1.641	1.721	4.333

#1	.2622	.2902	477.6	20.24	.8314
#2	.2180	.3116	464.4	19.46	.8329
#3	.3107	.2799	479.6	20.34	.9154
#4	.2457	.2736	484.8	19.89	.8473
#5	.2627	.2710	481.6	20.04	.8894

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	552.9	496.4	13.55	.3473	.2685
Stddev	2.2	3.7	.14	.0209	.0173
%RSD	.3935	.7442	1.028	6.032	6.446

#1	551.3	493.3	13.36	.3239	.2818
#2	555.5	493.2	13.56	.3787	.2879
#3	550.4	499.7	13.68	.3471	.2702
#4	552.6	500.9	13.48	.3336	.2568
#5	554.7	494.6	13.70	.3530	.2460

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	536.1	.2813
Stddev	4.9	.0223
%RSD	.9152	7.931

#1	536.8	.2608
#2	533.1	.2809
#3	538.6	.2576
#4	529.7	.3015
#5	542.4	.3057

Sample Name: ccv Run Time: 02/28/99 12:07

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	238.61
Stddev	.80
%RSD	.33601

#1	237.75
#2	238.52
#3	238.14
#4	238.79
#5	239.86

Method: ACTINIDE Sample Name: ccb

Operator: kjt

Comment: 2% hno3

Run Time: 02/28/99 12:20 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2499	.2429	.3391	.3032	.2535
Stddev	.0218	.0153	.0125	.0140	.0137
%RSD	8.724	6.307	3.690	4.612	5.419

#1	.2337	.2365	.3499	.2981	.2547
#2	.2448	.2225	.3477	.3018	.2684
#3	.2345	.2640	.3427	.3005	.2317
#4	.2496	.2425	.3187	.2891	.2609
#5	.2870	.2488	.3365	.3266	.2517

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2673	.3630	.2755	.2536	.2628
Stddev	.0098	.0157	.0126	.0232	.0145
%RSD	3.666	4.324	4.583	9.134	5.503

#1	.2631	.3849	.2645	.2337	.2659
#2	.2740	.3643	.2600	.2448	.2503
#3	.2528	.3525	.2865	.2388	.2458
#4	.2778	.3441	.2877	.2595	.2721
#5	.2686	.3690	.2785	.2913	.2799

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2889	.2779
Stddev	.0256	.0275
%RSD	8.861	9.906

#1	.2519	.2645
#2	.2837	.2934
#3	.3216	.3188
#4	.2862	.2552
#5	.3012	.2573

Analysis Report

Sample Name: ccb Run Time: 02/28/99 12:20

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	237.47
Stddev	1.55
%RSD	.65115

#1	238.17
#2	239.70
#3	237.35
#4	236.39
#5	235.76

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 10 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 3070 THRU 3400)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	INDUCTIVELY COUPLED PLASMA/MASS SPECTROMETER ANALYSIS, ICP/MS WORKLIST 27947 THRU 28912		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)

IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 27947

Analyst: bjg

Instrument: ICPMS1

Book#: 8859 990303-3

Method: LA-506-101 Rev/Mod

A-Ø

11 B59

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	ØMSAC-QC TH232	QC	0.0050	5.11e-03	102.200 % Recovery
1	ICV	0	ØMSAC-QC NP237	QC	4.94e-3	5.94e-03	120.243 % Recovery
1	ICV	0	ØMSAC-QC PU239	QC	0.0050	5.20e-03	104.000 % Recovery
1	ICV	0	ØMSAC-QC 241	QC	0.00513	5.65e-03	110.136 % Recovery
2	ICB	0	ØMSAC-QC PA231	QC	1	<5.57e-6	ug/mL
2	ICB	0	ØMSAC-QC TH229	QC	1	<4.46e-6	ug/mL
2	ICB	0	ØMSAC-QC TH230	QC	1	<3.55e-6	ug/mL
2	ICB	0	ØMSAC-QC TH232	QC	1	<1.93e-6	ug/mL
2	ICB	0	ØMSAC-QC NP237	QC	1	<4.67e-6	ug/mL
2	ICB	0	ØMSAC-QC PU239	QC	1	<3.47e-6	ug/mL
2	ICB	0	ØMSAC-QC PU240	QC	1	<3.55e-6	ug/mL
2	ICB	0	ØMSAC-QC PU242	QC	1	<2.78e-6	ug/mL
2	ICB	0	ØMSAC-QC 241	QC	1	<3.71e-6	ug/mL
2	ICB	0	ØMSAC-QC 243	QC	1	<3.83e-6	ug/mL
2	ICB	0	ØMSAC-QC 244	QC	1	<3.41e-6	ug/mL
3	BLNK-PREP	0	ØMSACTAL PA231-A1	SOLID	1	<5.57e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL TH229-A1	SOLID	1	<4.46e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL TH230-A1	SOLID	1	<3.55e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL TH232-A1	SOLID	1	<1.93e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL NP237-A1	SOLID	1	<4.67e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL PU239-A1	SOLID	1	<3.47e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL PU240-A1	SOLID	1	<3.55e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL PU242-A1	SOLID	1	<2.78e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL 241-A1	SOLID	1	<3.71e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL 243-A1	SOLID	1	<3.83e-5	ug/g
3	BLNK-PREP	0	ØMSACTAL 244-A1	SOLID	1	<3.41e-5	ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL PA231-A1	SOLID	N/A < 2.174e-01	0.217 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL TH229-A1	SOLID	N/A < 1.742e-01	0.174 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL TH230-A1	SOLID	N/A < 1.386e-01	0.139 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL TH232-A1	SOLID	N/A < 1.456e+00	7.53e-002 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL NP237-A1	SOLID	N/A < 1.822e-01	0.182 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL PU239-A1	SOLID	N/A < 1.354e-01	0.135 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL PU240-A1	SOLID	N/A < 1.387e-01	0.139 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL PU242-A1	SOLID	N/A < 1.085e-01	0.108 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL 241-A1	SOLID	N/A < 1.446e-01	0.145 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL 243-A1	SOLID	N/A < 1.497e-01	0.150 ug/g
4	SAMPLE	S98T002607	0 A	ØMSACTAL 244-A1	SOLID	N/A < 1.333e-01	0.133 ug/g
5	DUP	S98T002607	0 A	ØMSACTAL PA231-A1	SOLID	<2.17e-1	<2.24e-1 RPD
5	DUP	S98T002607	0 A	ØMSACTAL TH229-A1	SOLID	<1.74e-1	<1.79e-1 RPD
5	DUP	S98T002607	0 A	ØMSACTAL TH230-A1	SOLID	<1.39e-1	<1.43e-1 RPD
5	DUP	S98T002607	0 A	ØMSACTAL TH232-A1	SOLID	1.46e+00	1.36e+00 7.092 RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27947

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
5 DUP	S98T002607	0 A	OMSACT1 NF237-A1	SOLID	<1.82e-1	<1.87e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 PU239-A1	SOLID	<1.35e-1	<1.39e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 PU240-A1	SOLID	<1.39e-1	<1.43e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 PU242-A1	SOLID	<1.08e-1	<1.12e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 241-A1	SOLID	<1.45e-1	<1.49e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 243-A1	SOLID	<1.50e-1	<1.54e-1		RPD
5 DUP	S98T002607	0 A	OMSACT1 244-A1	SOLID	<1.33e-1	<1.37e-1		RPD
6 SAMPLE	S98T002608	0 A	OMSACT1 PA231-A1	SOLID	N/A	< 2.236e+01	22.370	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 TH229-A1	SOLID	N/A	< 1.792e+01	17.920	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 TH230-A1	SOLID	N/A	< 1.426e+01	14.260	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 TH232-A1	SOLID	N/A	< 7.750e+00	7.751	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 NP237-A1	SOLID	N/A	< 1.875e+01	18.750	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 PU239-A1	SOLID	N/A	< 1.393e+01	13.930	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 PU240-A1	SOLID	N/A	< 1.427e+01	14.270	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 PU242-A1	SOLID	N/A	< 1.116e+01	11.160	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 241-A1	SOLID	N/A	< 1.488e+01	14.880	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 243-A1	SOLID	N/A	< 1.540e+01	15.400	ug/g
6 SAMPLE	S98T002608	0 A	OMSACT1 244-A1	SOLID	N/A	< 1.371e+01	13.710	ug/g
7 DUP	S98T002608	0 A	OMSACT1 PA231-A1	SOLID	<2.24e1	<2.19e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 TH229-A1	SOLID	<1.79e1	<1.75e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 TH230-A1	SOLID	<1.43e1	<1.39e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 TH232-A1	SOLID	<7.75e0	<7.58e0		RPD
7 DUP	S98T002608	0 A	OMSACT1 NP237-A1	SOLID	<1.88e1	<1.83e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 PU239-A1	SOLID	<1.39e1	<1.36e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 PU240-A1	SOLID	<1.43e1	<1.40e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 PU242-A1	SOLID	<1.12e1	<1.09e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 241-A1	SOLID	<1.49e1	<1.46e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 243-A1	SOLID	<1.54e1	<1.51e1		RPD
7 DUP	S98T002608	0 A	OMSACT1 244-A1	SOLID	<1.37e1	<1.34e1		RPD
8 SPK-POST	S98T002608	0 A	OMSACT1 TH232-A1	SOLID	0.020	1.98e-02	99.000	% Recovery
8 SPK-POST	S98T002608	0 A	OMSACT1 NP237-A1	SOLID	0.01976	2.29e-02	115.891	% Recovery
8 SPK-POST	S98T002608	0 A	OMSACT1 PU239-A1	SOLID	0.02008	2.04e-02	101.594	% Recovery
8 SPK-POST	S98T002608	0 A	OMSACT1 241-A1	SOLID	0.02052	2.20e-02	107.212	% Recovery
9 SAMPLE	S98T002609	0 A	OMSACT1 PA231-A1	SOLID	N/A	< 2.300e+01	23.000	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 TH229-A1	SOLID	N/A	< 1.843e+01	18.430	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 TH230-A1	SOLID	N/A	< 1.467e+01	14.670	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 TH232-A1	SOLID	N/A	< 7.971e+00	7.970	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 NP237-A1	SOLID	N/A	< 1.928e+01	19.280	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 PU239-A1	SOLID	N/A	< 1.433e+01	14.330	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 PU240-A1	SOLID	N/A	< 1.467e+01	14.670	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 PU242-A1	SOLID	N/A	< 1.148e+01	11.480	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 241-A1	SOLID	N/A	< 1.530e+01	15.300	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 243-A1	SOLID	N/A	< 1.583e+01	15.830	ug/g
9 SAMPLE	S98T002609	0 A	OMSACT1 244-A1	SOLID	N/A	< 1.410e+01	14.100	ug/g
10 DUP	S98T002609	0 A	OMSACT1 PA231-A1	SOLID	<2.30e1	<2.16e1		RPD
10 DUP	S98T002609	0 A	OMSACT1 TH229-A1	SOLID	<1.84e1	<1.73e1		RPD
10 DUP	S98T002609	0 A	OMSACT1 TH230-A1	SOLID	<1.47e1	<1.38e1		RPD
10 DUP	S98T002609	0 A	OMSACT1 TH232-A1	SOLID	<7.97e0	<7.49e0		RPD
10 DUP	S98T002609	0 A	OMSACT1 NP237-A1	SOLID	<1.93e1	<1.81e1		RPD

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27947

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
10	DUP	S98T002609	0	A	@MSACTA1 PU239-A1	SOLID	<1.43e1	<1.35e1		RPD
10	DUP	S98T002609	0	A	@MSACTA1 PU240-A1	SOLID	<1.47e1	<1.38e1		RPD
10	DUP	S98T002609	0	A	@MSACTA1 PU242-A1	SOLID	<1.15e1	<1.08e1		RPD
10	DUP	S98T002609	0	A	@MSACTA1 241-A1	SOLID	<1.53e1	<1.44e1		RPD
10	DUP	S98T002609	0	A	@MSACTA1 243-A1	SOLID	<1.58e1	<1.49e1		RPD
10	DUP	S98T002609	0	A	@MSACTA1 244-A1	SOLID	<1.41e1	<1.33e1		RPD
11	CCV	0			@MSAC-QC TH232	QC	0.5e-2	5.15e-03	103.000 % Recovery	
11	CCV	0			@MSAC-QC NP237	QC	4.94e-3	6.08e-03	123.077 % Recovery	
11	CCV	0			@MSAC-QC PU239	QC	0.5e-2	5.38e-03	107.600 % Recovery	
11	CCV	0			@MSAC-QC 241	QC	0.513e-2	5.79e-03	112.865 % Recovery	
12	CCB	0			@MSAC-QC PA231	QC	1	<5.57e-6		ug/mL
12	CCB	0			@MSAC-QC TH229	QC	1	<4.46e-6		ug/mL
12	CCB	0			@MSAC-QC TH230	QC	1	<3.55e-6		ug/mL
12	CCB	0			@MSAC-QC TH232	QC	1	<1.93e-6		ug/mL
12	CCB	0			@MSAC-QC NP237	QC	1	<4.67e-6		ug/mL
12	CCB	0			@MSAC-QC PU239	QC	1	<3.47e-6		ug/mL
12	CCB	0			@MSAC-QC PU240	QC	1	<3.55e-6		ug/mL
12	CCB	0			@MSAC-QC PU242	QC	1	<2.78e-6		ug/mL
12	CCB	0			@MSAC-QC 241	QC	1	<3.71e-6		ug/mL
12	CCB	0			@MSAC-QC 243	QC	1	<3.83e-6		ug/mL
12	CCB	0			@MSAC-QC 244	QC	1	<3.41e-6		ug/mL

Final page for worklist# 27947

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

H. Nastos
Reviewer Signature _____ Date 3-4-99

Np- 237 ICV = 120.2%, CCV 123%. Will
be rerun if requested.

01/21/99 12:43

A-0004-1

FILE # AC0803A.7X5

Page: 1

LABCORE Data Entry Template for Worklist# 27947

Analyst: BTG Instrument: ICPMS1 Book# 8858, 990303-3 (11858)Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			MSAC-QC	QC		
2 ICB			MSAC-QC	QC		
3 BLNK-PREP			MSACTA1	SOLID		
4 SAMPLE	S98T002607 0 A		MSACTA1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1, 243-A1, 244-A1, NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
5 DUP	S98T002607 0 A		MSACTA1	SOLID		
6 SAMPLE	S98T002608 0 A		MSACTA1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1, 243-A1, 244-A1, NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
7 DUP	S98T002608 0 A		MSACTA1	SOLID		
8 SPK-POST	S98T002608 0 A		MSACTA1	SOLID		
9 SAMPLE	S98T002609 0 A		MSACTA1	SOLID	98000436	BY-105 (2)
Analytes Requested: 241-A1, 243-A1, 244-A1, NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
10 DUP	S98T002609 0 A		MSACTA1	SOLID		
11 CCV			MSAC-QC	QC		
12 CCB			MSAC-QC	QC		

Final page for worklist # 27947

Analyst Signature BTG Date 3-3-99

Analyst Signature _____ Date _____

PREPARE 1st - 9-1
 S98T002607 .05-1-10-1
 2607-D .05-1-10-1
 S98T002608 .05-1-10-1, 1st-1-10-1
 2608-A
 2608-A
 S98T002609
 2609-D

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

- A POST-SAMPLE = .04 ml of MSAC. 50% 8858, 10-1.

SCANS = 982607, 982608, 982609

* DILUTED MSAC ICV 1:1 ... 990303-3 (11858)

1000

3073

XAF 80:02 88/03/00

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/03/99 17:42 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	227000.
Stddev	2025.
%RSD	.8919

#1	229800.
#2	226400.
#3	225700.
#4	229000.
#5	228600.
#6	224800.
#7	225700.
#8	229500.
#9	226000.
#10	224500.

bjg 3-3-99
By-105 AND DIRECT
538700 2607, 08, 09
FILE# AC0303A.TTX
W.L.# 27947

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3074 TO 3104
SKW 3/24/99

Method: ACTINIDE Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/03/99 17:43 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3471	.3025	.4786	.3953	.3470
Stddev	.0426	.0272	.0411	.0486	.0563
%RSD	12.27	9.001	8.584	12.29	16.23

#1	.3602	.2825	.5154	.3881	.2950
#2	.2774	.3232	.4390	.3902	.3232
#3	.3789	.3388	.5205	.4251	.3327
#4	.3383	.2926	.4846	.4511	.3414
#5	.3809	.2756	.4335	.3220	.4428

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3591	.5451	.3897	.3377	.3630
Stddev	.0181	.0516	.0447	.0279	.0225
%RSD	5.046	9.467	11.47	8.256	6.205

#1	.3353	.6272	.3508	.3322	.4005
#2	.3811	.5427	.4024	.2927	.3658
#3	.3635	.5452	.4127	.3511	.3450
#4	.3688	.5243	.4450	.3658	.3475
#5	.3468	.4861	.3375	.3468	.3561

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3500	.3432
Stddev	.0342	.0264
%RSD	9.768	7.698

#1	.3974	.3291
#2	.3750	.3719
#3	.3234	.3573
#4	.3322	.3048
#5	.3220	.3530

Sample Name: blank Run Time: 03/03/99 17:43

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	108.39
Stddev	.93
%RSD	.85994

#1	107.36
#2	109.34
#3	108.22
#4	109.36
#5	107.65

3076

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 17:48 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3463	.3815	406.2	7.302	.8848
Stddev	.0444	.0430	8.1	.186	.0872
%RSD	12.81	11.26	1.990	2.548	9.859
#1	.3974	.4337	392.8	7.010	.7644
#2	.2924	.4137	408.2	7.353	1.002
#3	.3842	.3370	412.5	7.441	.8503
#4	.3164	.3822	412.2	7.240	.9179
#5	.3411	.3411	405.2	7.467	.8900

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	513.1	463.8	13.17	.4240	.3595
Stddev	11.6	12.7	.42	.0434	.0343
%RSD	2.269	2.745	3.197	10.22	9.551
#1	493.7	442.0	12.51	.4610	.3913
#2	513.6	463.6	13.30	.4043	.3141
#3	522.5	468.9	13.01	.4787	.3842
#4	521.9	473.7	13.57	.3979	.3759
#5	513.6	470.9	13.43	.3783	.3318

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	519.2	.3898
Stddev	9.6	.0259
%RSD	1.839	6.652
#1	503.3	.4247
#2	518.0	.4074
#3	526.8	.3779
#4	521.8	.3791
#5	526.1	.3597

Sample Name: icv Run Time: 03/03/99 17:48

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	107.36
Stdddev	1.56
%RSD	1.4491
#1	109.89
#2	107.17
#3	105.85
#4	106.40
#5	107.50

Method: ACTINIDE Sample Name: ich

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 17:57 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3087	.3326	.4777	.3673	.3089
Stddev	.0563	.0304	.0287	.0214	.0390
%RSD	18.24	9.125	6.014	5.830	12.62

#1	.2822	.2790	.4946	.3773	.2727
#2	.2440	.3485	.4911	.3612	.3263
#3	.3651	.3432	.5086	.3994	.3682
#4	.3709	.3397	.4457	.3522	.2805
#5	.2814	.3526	.4484	.3464	.2969

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3177	.4838	.3452	.3418	.3534
Stddev	.0159	.0432	.0301	.0330	.0254
%RSD	4.997	8.924	8.722	9.652	7.200

#1	.3329	.4756	.3488	.3583	.3329
#2	.3073	.4911	.3168	.3834	.3548
#3	.3058	.5523	.3900	.3370	.3963
#4	.3054	.4364	.3179	.3366	.3366
#5	.3371	.4639	.3526	.2938	.3464

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3288	.3276
Stddev	.0400	.0259
%RSD	12.15	7.900

#1	.2663	.2949
#2	.3295	.3232
#3	.3495	.3620
#4	.3740	.3148
#5	.3247	.3433

3079

Sample Name: icb Run Time: 03/03/99 17:57

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	106.38
Stddev	1.17
%RSD	1.0957
#1	105.13
#2	105.21
#3	106.83
#4	106.95
#5	107.78

Method: ACTINIDE . Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/03/99 18:01 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3067	.2996	.5456	.6087	.3053
Stddev	.0216	.0353	.0254	.0605	.0290
%RSD	7.044	11.79	4.648	9.933	9.508

#1	.3385	.2806	.5872	.7001	.2748
#2	.2844	.2815	.5372	.5717	.3045
#3	.2895	.2614	.5425	.5593	.3457
#4	.3065	.3375	.5428	.6412	.2812
#5	.3148	.3371	.5182	.5711	.3204

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2931	.6826	.3488	.3448	.3360
Stddev	.0172	.0812	.0401	.0241	.0448
%RSD	5.876	11.89	11.50	6.984	13.34

#1	.3037	.8187	.3645	.3790	.3992
#2	.2988	.6464	.3763	.3246	.3246
#3	.2698	.6184	.3148	.3598	.3289
#4	.2812	.6946	.2981	.3375	.3515
#5	.3120	.6352	.3900	.3232	.2758

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3374	.3317
Stddev	.0362	.0258
%RSD	10.73	7.768

#1	.3182	.3153
#2	.3821	.3562
#3	.3710	.3626
#4	.3065	.3065
#5	.3093	.3176

3081

Sample Name: prepblk Run Time: 03/03/99 18:01

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	117.61
Stddev	1.88
%RSD	1.5964
#1	115.23
#2	116.03
#3	118.59
#4	118.53
#5	119.64

Analysis Report

Method: ACTINIDE Sample Name: s98t002607

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 18:23 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2957	.2917	3.436	128.9	.2847
Stddev	.0344	.0282	.081	1.8	.0235
%RSD	11.64	9.658	2.358	1.400	8.268
#1	.2765	.2894	3.447	128.0	.3119
#2	.2952	.2732	3.549	126.7	.2607
#3	.3254	.2899	3.470	130.6	.3060
#4	.2493	.3388	3.359	130.9	.2812
#5	.3322	.2670	3.356	128.3	.2639

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5417	.5737	.3209	.3291	.3304
Stddev	.0405	.0507	.0310	.0398	.0387
%RSD	7.483	8.845	9.651	12.09	11.71
#1	.5370	.5884	.3280	.3633	.3923
#2	.5339	.5747	.2889	.3486	.2952
#3	.6024	.6475	.3705	.3608	.3350
#4	.4890	.5113	.3068	.2780	.3004
#5	.5464	.5464	.3105	.2950	.3291

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3369	.2774
Stddev	.0254	.0329
%RSD	7.540	11.85
#1	.3376	.2347
#2	.3392	.2889
#3	.3060	.3222
#4	.3260	.2589
#5	.3757	.2825

Analysis Report

03/03/99 18:26:54

page 2

Sample Name: s98t002607 Run Time: 03/03/99 18:23

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.99
Stddev	1.69
%RSD	1.6128

#1	103.67
#2	106.14
#3	103.47
#4	104.30
#5	107.36

3084

Method: ACTINIDE Sample Name: s98t002607_d

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 18:28 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3058	.3011	3.171	116.5	.2770
Stddev	.0454	.0384	.132	2.3	.0441
%RSD	14.86	12.74	4.152	1.996	15.92

#1	.2776	.2898	3.176	116.0	.2837
#2	.3398	.2947	3.004	118.5	.2286
#3	.2777	.3493	3.353	114.7	.2687
#4	.2657	.3240	3.097	113.9	.2569
#5	.3683	.2475	3.227	119.3	.3471

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4950	.5006	.2930	.2696	.2965
Stddev	.0501	.0336	.0288	.0417	.0240
%RSD	10.12	6.705	9.832	15.47	8.099

#1	.4912	.5522	.2929	.2288	.2624
#2	.5684	.4662	.2586	.3398	.3158
#3	.4269	.5016	.3374	.2538	.2806
#4	.4934	.4758	.2802	.2598	.3065
#5	.4950	.5071	.2958	.2656	.3169

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2965	.3047
Stddev	.0491	.0262
%RSD	16.55	8.581

#1	.2258	.2593
#2	.3068	.3128
#3	.3284	.3075
#4	.2715	.3211
#5	.3501	.3230

Sample Name: s98t002607_d Run Time: 03/03/99 18:28

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.27
Stddev	1.84
%RSD	1.6560

#1	109.26
#2	110.83
#3	111.65
#4	114.18
#5	110.43

Method: ACTINIDE Sample Name: s98t002608

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 18:32 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2793	.2737	.3868	879.7	.2728
Stddev	.0127	.0392	.0451	5.4	.0091
%RSD	4.545	14.33	11.67	.6170	3.317

#1	.2614	.2643	.3920	873.2	.2846
#2	.2861	.3245	.4543	877.2	.2625
#3	.2918	.2254	.3438	877.4	.2774
#4	.2863	.3009	.3973	885.7	.2746
#5	.2709	.2534	.3467	885.0	.2651

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.761	.6391	.3033	.2822	.2974
Stddev	.034	.0484	.0337	.0187	.0177
%RSD	1.904	7.571	11.11	6.614	5.950

#1	1.769	.5576	.2701	.2759	.2962
#2	1.720	.6608	.3392	.2743	.2832
#3	1.786	.6790	.3120	.2918	.2774
#4	1.733	.6340	.2659	.2600	.3185
#5	1.797	.6642	.3292	.3088	.3117

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2774	.3090
Stddev	.0310	.0198
%RSD	11.19	6.399

#1	.2846	.3340
#2	.2271	.2891
#3	.2831	.3178
#4	.3126	.3155
#5	.2797	.2884

3087

Sample Name: s98t002608 Run Time: 03/03/99 18:32

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	114.33
Stddev	.89
%RSD	.77430

#1	114.78
#2	113.00
#3	115.37
#4	114.09
#5	114.42

Method: ACTINIDE Sample Name: s98t002608_d Operator: bjg
 Comment: 20301 df
 Run Time: 03/03/99 18:37 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2854	.2860	.3642	773.3	.2838
Stddev	.0218	.0201	.0367	7.6	.0076
%RSD	7.647	7.036	10.09	.9823	2.673
#1	.2736	.3081	.4002	770.1	.2793
#2	.2627	.2740	.3729	771.2	.2825
#3	.2741	.2683	.3087	785.8	.2943
#4	.3022	.3079	.3905	774.0	.2879
#5	.3144	.2719	.3484	765.6	.2748

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.515	.6331	.2911	.2836	.3060
Stddev	.052	.0405	.0320	.0241	.0194
%RSD	3.407	6.403	11.00	8.502	6.357
#1	1.503	.6076	.2937	.2793	.3254
#2	1.461	.6837	.3418	.3221	.2966
#3	1.512	.6578	.2770	.2741	.2770
#4	1.499	.6357	.2879	.2565	.3193
#5	1.600	.5807	.2549	.2861	.3116

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3001	.2895
Stddev	.0250	.0349
%RSD	8.345	12.06
#1	.2937	.3081
#2	.3334	.2712
#3	.2856	.3145
#4	.2708	.2366
#5	.3173	.3173

Sample Name: s98t002608_d Run Time: 03/03/99 18:37

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	116.78
Stddev	1.11
%RSD	.94631

#1	115.76
#2	117.99
#3	115.54
#4	116.94
#5	117.68

Method: ACTINIDE Sample Name: s98t002608_a

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 18:42 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3138	.4577	1573.	916.0	2.426
Stddev	.0191	.0322	15.	6.4	.047
%RSD	6.096	7.036	.9293	.7024	1.931
#1	.3106	.4289	1563.	927.2	2.435
#2	.2906	.4874	1575.	912.7	2.393
#3	.3254	.4970	1598.	911.5	2.479
#4	.3396	.4311	1568.	915.9	2.365
#5	.3029	.4440	1562.	912.9	2.458

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1975.	1815.	49.37	.5621	.2949
Stddev	8.	28.	.39	.0302	.0293
%RSD	.3813	1.542	.7816	5.371	9.952
#1	1972.	1804.	49.16	.5384	.3165
#2	1978.	1791.	49.67	.5255	.3259
#3	1979.	1856.	49.21	.6006	.2544
#4	1983.	1792.	48.95	.5699	.3012
#5	1964.	1831.	49.88	.5764	.2764

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	2021.	.3273
Stddev	25.	.0310
%RSD	1.261	9.467
#1	1983.	.3431
#2	2013.	.3053
#3	2046.	.2870
#4	2043.	.3366
#5	2021.	.3646

Sample Name: s98t002608_a Run Time: 03/03/99 18:42

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	113.03
Stddev	.40
%RSD	.35150

#1	112.68
#2	113.54
#3	112.67
#4	112.89
#5	113.35

Method: ACTINIDE Sample Name: s98t002609

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 18:52 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2818	.2941	.5073	651.0	.3040
Stddev	.0277	.0408	.0688	13.3	.0115
%RSD	9.834	13.86	13.56	2.038	3.773
#1	.2888	.3064	.5485	634.3	.2947
#2	.2930	.3379	.4545	658.8	.3169
#3	.2827	.2496	.6045	656.8	.3038
#4	.3089	.3240	.4906	665.2	.2907
#5	.2353	.2527	.4387	639.7	.3138

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.354	1.201	.3089	.3107	.3258
Stddev	.091	.112	.0178	.0268	.0307
%RSD	6.758	9.296	5.767	8.623	9.425
#1	1.319	1.234	.3093	.2888	.2801
#2	1.393	1.256	.3020	.3259	.3319
#3	1.401	1.260	.3007	.2797	.3609
#4	1.211	1.254	.3392	.3452	.3422
#5	1.444	1.002	.2934	.3138	.3138

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2970	.2968
Stddev	.0181	.0277
%RSD	6.101	9.339
#1	.3093	.3326
#2	.2811	.2781
#3	.2947	.2737
#4	.3210	.3210
#5	.2789	.2789

3093

Sample Name: s98t002609 Run Time: 03/03/99 18:52

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	112.28
Stddev	2.09
%RSD	1.8629

#1	114.25
#2	111.48
#3	110.84
#4	110.08
#5	114.74

Method: ACTINIDE Sample Name: s98t002609-^D
 Comment: 20301 df ^{5/4 3-3-99}
 Run Time: 03/03/99 18:57 Type: Unk Mode: IR Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2978	.3072	.4209	888.3	.2999
Stddev	.0376	.0271	.0454	17.2	.0119
%RSD	12.64	8.818	10.77	1.936	3.957

#1	.3272	.2618	.4343	867.2	.3094
#2	.2548	.3058	.4257	888.6	.2968
#3	.3115	.3324	.4702	882.3	.2815
#4	.2606	.3188	.3464	914.8	.3005
#5	.3350	.3171	.4278	888.7	.3111

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.772	1.162	.3291	.3308	.2961
Stddev	.088	.038	.0551	.0632	.0405
%RSD	4.964	3.260	16.75	19.10	13.66

#1	1.714	1.169	.2737	.3183	.2707
#2	1.703	1.148	.2788	.3028	.3328
#3	1.746	1.201	.3444	.3414	.2695
#4	1.778	1.186	.4078	.2606	.2606
#5	1.921	1.104	.3410	.4308	.3470

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3230	.2998
Stddev	.0581	.0222
%RSD	18.00	7.412

#1	.3481	.2737
#2	.4048	.3238
#3	.3205	.3145
#4	.2514	.2790
#5	.2902	.3081

Sample Name: s98t002602 ⁴ Run Time: 03/03/99 18:57

98-33-98

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	110.94
Stddev	1.28
%RSD	1.1548

#1	112.05
#2	111.18
#3	111.30
#4	108.72
#5	111.43

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/03/99 19:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3056	.3625	409.7	7.586	.8933
Stddev	.0173	.0191	6.7	.246	.0314
%RSD	5.676	5.272	1.637	3.242	3.519
#1	.3090	.3708	400.3	7.319	.8623
#2	.3160	.3586	410.4	7.767	.8995
#3	.2905	.3913	410.5	7.692	.8804
#4	.3268	.3478	419.0	7.828	.9444
#5	.2854	.3443	408.0	7.324	.8799

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	525.4	480.1	13.23	.3685	.3328
Stddev	6.4	11.8	.39	.0119	.0330
%RSD	1.225	2.461	2.909	3.239	9.925
#1	522.6	470.4	12.96	.3708	.3826
#2	532.8	490.7	13.37	.3555	.3221
#3	521.8	475.7	12.82	.3587	.3290
#4	531.6	494.6	13.81	.3717	.3388
#5	518.2	469.0	13.22	.3855	.2913

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	531.9	.3295
Stddev	6.8	.0259
%RSD	1.276	7.866
#1	526.1	.3090
#2	541.6	.3707
#3	528.4	.3379
#4	536.2	.3208
#5	526.9	.3090

Sample Name: ccv Run Time: 03/03/99 19:02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.97
Stddev	1.53
%RSD	1.3683

#1	113.26
#2	109.70
#3	112.45
#4	111.19
#5	113.27

3098

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 19:10 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3274	.2857	.4846	.3742	.2975
Stddev	.0271	.0219	.0163	.0214	.0199
%RSD	8.287	7.665	3.373	5.728	6.680
#1	.3155	.2972	.4902	.3891	.3033
#2	.3325	.2726	.4823	.3684	.2965
#3	.3624	.2563	.4802	.3918	.3241
#4	.2892	.3128	.5076	.3394	.2951
#5	.3373	.2895	.4626	.3820	.2686

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3296	.5662	.3246	.3044	.3022
Stddev	.0377	.0596	.0344	.0268	.0253
%RSD	11.45	10.53	10.59	8.811	8.362
#1	.3217	.6250	.3431	.3492	.2849
#2	.2786	.4972	.2786	.2846	.2786
#3	.3388	.5067	.3300	.2828	.2917
#4	.3837	.5962	.3040	.3040	.3394
#5	.3253	.6059	.3671	.3015	.3164

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3279	.3093
Stddev	.0389	.0230
%RSD	11.87	7.440
#1	.3921	.2910
#2	.3055	.2786
#3	.3064	.3241
#4	.2981	.3306
#5	.3373	.3224

Analysis Report

03/03/99 19:14:31

P-3-

HNF-1663 REV. 0

Sample Name: ccb Run Time: 03/03/99 19:10

Int. Std.	1x(1931
Line	193/pulse
Units	Cts/S
Avg	111.57
Stddev	1.74
%RSD	1.5569

#1	108.81
#2	111.28
#3	113.14
#4	112.94
#5	111.68

3100

LABCORE Completed Worklist Report for Worklist# 28434

Analyst: bjg

Instrument: ICPMS1

Book#: 99D301-1

Method: LA-506-101 Rev/Mod A-2

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000142	1.48e-04	104.225 % Recovery	
1	ICV	0	QMSU-QC U238	QC	0.020	2.13e-02	106.500 % Recovery	
1	ICB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	STD-PREP	0	QMSU-A1 U238-A1	SOLID	10	1.04e+01	104.000 % Recovery	
4	BLNK-PREP	0	QMSU-A1 U233-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U234-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U235-A1	SOLID	1	<1.20e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U236-A1	SOLID	1	<1.60e-4	ug/g	
4	BLNK-PREP	0	QMSU-A1 U238-A1	SOLID	1	<1.20e-4	ug/g	
5	SAMPLE	S98T003399	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.886e+01	48.860 ug/g	
5	SAMPLE	S98T003399	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.886e+01	48.860 ug/g	
5	SAMPLE	S98T003399	0 A QMSU-A1 U235-A1	SOLID	N/A	< 2.622e+02	48.860 ug/g	
5	SAMPLE	S98T003399	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.515e+01	65.150 ug/g	
5	SAMPLE	S98T003399	0 A QMSU-A1 U238-A1	SOLID	N/A	< 4.100e+04	48.860 ug/g	
6	DUP	S98T003399	0 A QMSU-A1 U233-A1	SOLID	<4.89e1	<4.67e1	RPD	
6	DUP	S98T003399	0 A QMSU-A1 U234-A1	SOLID	<4.89e1	<4.67e1	RPD	
6	DUP	S98T003399	0 A QMSU-A1 U235-A1	SOLID	2.62e+02	3.02e+02	14.184 RPD	
6	DUP	S98T003399	0 A QMSU-A1 U236-A1	SOLID	<6.52e1	<6.23e1	RPD	
6	DUP	S98T003399	0 A QMSU-A1 U238-A1	SOLID	4.10e+04	4.79e+04	15.523 RPD	
7	SPK	S98T003399	0 A QMSU-A1 U235-A1	SOLID	14.22	7.64e+01	537.271 % Recovery	
7	SPK	S98T003399	0 A QMSU-A1 U238-A1	SOLID	2000	1.17e+04	585.000 % Recovery	
8	SPK-POST	S98T003399	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.35e-04	94.937 % Recovery	
8	SPK-POST	S98T003399	0 A QMSU-A1 U238-A1	SOLID	0.020	2.01e-02	100.500 % Recovery	
9	SAMPLE	S98T002930	0 A QMSU-A1 U233-A1	SOLID	N/A	< 4.689e+00	4.689 ug/g	
9	SAMPLE	S98T002930	0 A QMSU-A1 U234-A1	SOLID	N/A	< 4.689e+00	4.689 ug/g	
9	SAMPLE	S98T002930	0 A QMSU-A1 U235-A1	SOLID	N/A	< 1.191e+02	4.689 ug/g	
9	SAMPLE	S98T002930	0 A QMSU-A1 U236-A1	SOLID	N/A	< 6.252e+00	6.251 ug/g	
9	SAMPLE	S98T002930	0 A QMSU-A1 U238-A1	SOLID	N/A	< 1.837e+04	4.689 ug/g	
10	DUP	S98T002930	0 A QMSU-A1 U233-A1	SOLID	<4.69e0	<4.76e0	RPD	
10	DUP	S98T002930	0 A QMSU-A1 U234-A1	SOLID	<4.69e0	<4.76e0	RPD	
10	DUP	S98T002930	0 A QMSU-A1 U235-A1	SOLID	1.19e+02	7.42e+01	46.377 RPD	
10	DUP	S98T002930	0 A QMSU-A1 U236-A1	SOLID	<6.25e0	<6.35e0	RPD	
10	DUP	S98T002930	0 A QMSU-A1 U238-A1	SOLID	1.84e+04	1.18e+04	43.709 RPD	
11	SPK	S98T002930	0 A QMSU-A1 U235-A1	SOLID	14.22	-2.78e1	-195.499 % Recovery	
11	SPK	S98T002930	0 A QMSU-A1 U238-A1	SOLID	2000	-4.40e3	-220.000 % Recovery	
12	SPK-POST	S98T002930	0 A QMSU-A1 U235-A1	SOLID	0.0001422	1.37e-04	96.343 % Recovery	
12	SPK-POST	S98T002930	0 A QMSU-A1 U238-A1	SOLID	0.020	2.09e-02	104.500 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28434

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
13 CCV	0	QMSU-QC	U235	QC	0.000142	1.52e-04	107.042 % Recovery	
13 CCV	0	QMSU-QC	U238	QC	0.020	2.25e-02	112.500 % Recovery	
14 CCB	0	QMSU-QC	U233	QC	1	<1.20e-5		ug/mL
14 CCB	0	QMSU-QC	U234	QC	1	<1.20e-5		ug/mL
14 CCB	0	QMSU-QC	U235	QC	1	<1.20e-5		ug/mL
14 CCB	0	QMSU-QC	U236	QC	1	<1.60e-5		ug/mL
14 CCB	0	QMSU-QC	U238	QC	1	<1.20e-5		ug/mL

Final page for worklist# 28434

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

H. Anastos 3-3-99
Reviewer Signature Date

* See Attached for notes
on QC failures.

03/03/99 08:09

WS2

Page: 1

LABCORE Data Entry Template for Worklist# 28434

Analyst: BIG Instrument: ICPMS1 Book# 990301-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV			@MSU-QC	QC		
2	ICB			@MSU-QC	QC		
~~3~~	~~ILS~~			~~@MSU-QC~~	~~QC~~		~~deleted Ilea 3-3-99~~
4	STD-PREP			@MSU-A1	SOLID		
5	BLNK-PREP			@MSU-A1	SOLID		
6	SAMPLE	S98T003399	0	A	@MSU-A1	SOLID	98000436 BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
7	DUP	S98T003399	0	A	@MSU-A1	SOLID	
8	SPK	S98T003399	0	A	@MSU-A1	SOLID	
9	SPK-POST	S98T003399	0	A	@MSU-A1	SOLID	
10	SAMPLE	S98T002930	0	A	@MSU-A1	SOLID	98000490 BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
11	DUP	S98T002930	0	A	@MSU-A1	SOLID	
12	SPK	S98T002930	0	A	@MSU-A1	SOLID	
13	SPK-POST	S98T002930	0	A	@MSU-A1	SOLID	
14	CCV			@MSU-QC	QC		
15	CCB			@MSU-QC	QC		

Final page for worklist # 28434

Greche 3-3-99

Data Entry Comments:

Re-dig Spk failure on S98T003399 due to high
level of U in sample. Same for S98T002930.
Post spks all passed.

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 28434

S Type	Sample#	R A	Test	Matrix	Group#	Project	Date
Signature		Date			Signature		
PREP STD			05-1-10-1		DF	201	
PREP BLK			1-1-8-1			10	
598700	3399		05-1-10-1-10-1-10-1			20,351	
	3399-D						
	3399-A						
	3399-S						
598700	2930		05-1-10-1-1-1-8-1			2010	
	2930-D						
	2930-A						
	2930-S						

SCANS = 982930d, 983399

POST SPIKES - .02 ml of 10ppm USTd in 10-1 (20ppm)

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3104

03/03/99 08:09

WS2

Page: 1

LABCORE Data Entry Template for Worklist# 28434

Analyst: BJG Instrument: ICPMS1 _____ Book# 990301-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV			@MSU-QC	QC		
2	ICB			@MSU-QC	QC		
3	LLS			@MSU-QC	QC		
4	STD-PREP			@MSU-A1	SOLID		
5	BLNK-PREP			@MSU-A1	SOLID		
6	SAMPLE	S98T003399	0	A	@MSU-A1	SOLID	98000436 BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
7	DUP	S98T003399	0	A	@MSU-A1	SOLID	
8	SPK	S98T003399	0	A	@MSU-A1	SOLID	
9	SPK-POST	S98T003399	0	A	@MSU-A1	SOLID	
10	SAMPLE	S98T002930	0	A	@MSU-A1	SOLID	98000490 BY-105 (2)
Analytes Requested: U233-A1 , U234-A1 , U235-A1 , U236-A1 , U238-A1							
11	DUP	S98T002930	0	A	@MSU-A1	SOLID	
12	SPK	S98T002930	0	A	@MSU-A1	SOLID	
13	SPK-POST	S98T002930	0	A	@MSU-A1	SOLID	
14	CCV			@MSU-QC	QC		
15	CCB			@MSU-QC	QC		

Final page for worklist # 28434

BJG 3-3-99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3105

ws2

LABCORE Data Entry Template for Worklist# 28434

S Type	Sample#	R A	Test	Matrix	Group#	Project	Date
Signature		Date			Signature		
					DF		
PREP STD				05-1-10-1		201	
PREP BLK				1-1-8-1		10	
S 98700 3399				05-1-10-1-1-1-10-1		20,351	
3399-D							
3399-A							
3399-S							
S 98700 2930				05-1-10-1-1-1-8-1		2010	
2930-D							
2930-A							
2930-S							

SCANS = 982930d, 983399

POST SPIKES - .02 ml of 10ppm U STD in 10-1 (20 ppb)

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3106

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/03/99

10:34 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/s
Avg	200000.
Stddev	1625.
%RSD	.8124

#1	197100.
#2	201800.
#3	199100.
#4	201900.
#5	201700.
#6	199300.
#7	199200.
#8	199000.
#9	199400.
#10	201600.

Bjg
3-3-99

BY-101 ACID DRIFT
598T00 3399, 2930
FILE # U0303A.TAT
W.L. # 28434

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3101 TO 3128.

HNF-1663 REV. 0

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 03/03/99 10:36 Type: Std

Mode: IR

Corr. Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.192	4.490	4.393	4.607	5.203
Stddev	.492	.734	.505	.632	.464
%RSD	11.75	16.35	11.51	13.73	8.913

#1	3.905	3.849	4.018	4.697	4.811
#2	3.868	4.653	5.046	4.373	4.990
#3	4.298	3.619	4.524	3.676	4.806
#4	5.012	5.012	4.609	5.358	5.704
#5	3.878	5.318	3.767	4.930	5.706

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.8600
Stddev	.1275
%RSD	1.4387

#1	8.8346
#2	8.9185
#3	8.8423
#4	8.6783
#5	9.0261

3108

El Name	Slope	Y-int	Correlation	Date Stdized
U_[233]	916.0699	4.1922	1.0000000	03/03/99 10:38:05

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.1922
100ppb U	0.711	0.711	0	0	655.52

U [233]
0
(S) IR⁷³⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[234]	915.6508	4.4902	1.0000000	03/03/99 10:38:05

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4902
100ppb U	0.711	0.711	0	0	655.52

U [234]
0
(S) IR⁷³⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[235]	915.7879	4.3927	1.0000000	03/03/99 10:38:05

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.3927
100ppb U	0.711	0.711	0	0	655.52

U [235]
0
(S) IR⁷³⁰⁰¹⁰
Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[236]	915.4868	4.6068	1.0000000	03/03/99 10:38:05

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.6068
100ppb U	0.711	0.711	0	0	655.52

U [236]
0
(S) IR⁷³⁰⁰¹⁰
Concentration

3109

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	826.8319	5.2032	1.0000000	03/03/99 10:38:05

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc	%	(S)IR
Blank	0	0	0	0	5.2032
100ppb U	99.3	99.3	0	0	82110

U [238]

0

(S)IR

Concentration

3110

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 03/03/99 10:40 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	653.5	653.5	653.5	653.5	80540.
Stddev	10.4	10.4	10.4	10.4	1699.
%RSD	1.591	1.591	1.591	1.591	2.109

#1	665.5	665.5	665.5	665.5	78800.
#2	637.8	637.8	637.8	637.8	80980.
#3	658.1	658.1	658.1	658.1	82430.
#4	649.8	649.8	649.8	649.8	81750.
#5	656.1	656.1	656.1	656.1	78730.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.1052
Stddev	.0949
%RSD	1.0418

#1	9.0619
#2	9.1545
#3	8.9585
#4	9.1633
#5	9.1877

3111

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	913.1923	4.1922	1.0000000	03/03/99 10:41:17

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.1922
100ppb U	0.711	0.711	0	0	653.47

U [233]
0
(S) IR⁷²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	912.7731	4.4902	1.0000000	03/03/99 10:41:17

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4902
100ppb U	0.711	0.711	0	0	653.47

U [234]
0
(S) IR⁷²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	912.9102	4.3927	1.0000000	03/03/99 10:41:17

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.3927
100ppb U	0.711	0.711	0	0	653.47

U [235]
0
(S) IR⁷²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	912.6091	4.6068	1.0000000	03/03/99 10:41:17

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.6068
100ppb U	0.711	0.711	0	0	653.47

U [236]
0
(S) IR⁷²⁰⁰¹⁰

Concentration

3112

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	811.0219	5.2032	1.0000000	03/03/99 10:41:17

Standard	Concentration		Difference		Signal
Name	Stated	Found	Conc.	%	(S) IR
Blank	0	0	0	0	5.2032
100ppb U	99.3	99.3	0	0	80540

U [238]

0

(S) IR 885000IU

Concentration

~~HNF-1663 REV. 0~~
LAB 5-11-99

Method: ISOU

Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/03/99 10:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	-.0005	-.0001	-.0005	.0001
Stddev	.0007	.0005	.0005	.0004	.0005
%RSD	323.8	96.65	328.2	86.34	437.7
#1	.0000	-.0011	-.0010	.0002	.0000
#2	-.0014	.0003	.0000	-.0004	.0008
#3	.0002	-.0008	-.0001	-.0008	-.0005
#4	.0003	-.0007	.0001	-.0009	.0004
#5	-.0002	-.0004	.0002	-.0005	.0000

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.5768
Stddev	.0751
%RSD	.78427

#1	9.5456
#2	9.5734
#3	9.6585
#4	9.4701
#5	9.6361

3114

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/03/99 10:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	.0010	.1476	.0022	21.34
Stddev	.0007	.0006	.0040	.0006	.17
%RSD	180.6	67.78	2.728	27.09	.7734
#1	-.0014	.0005	.1453	.0025	21.13
#2	-.0003	.0008	.1442	.0031	21.21
#3	-.0004	.0002	.1544	.0014	21.39
#4	.0005	.0016	.1479	.0022	21.44
#5	-.0003	.0016	.1463	.0021	21.53

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.7443
Stddev	.0899
%RSD	.92225

#1	9.8861
#2	9.7181
#3	9.6945
#4	9.6528
#5	9.7699

Method: ISOU Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 11:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	-.0009	-.0004	-.0009	-.0010
Stddev	.000	.0006	.0008	.0006	.0005
%RSD	823.1	74.36	178.1	66.47	53.83

#1	-.0005	-.0015	-.0011	-.0017	-.0003
#2	-.0002	-.0015	.0000	-.0001	-.0012
#3	.0000	-.0001	.0007	-.0012	-.0014
#4	.0003	-.0009	-.0012	-.0011	-.0013
#5	.0002	-.0003	-.0006	-.0005	-.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	9.8425
Stddev	.1302
%RSD	1.3229

#1	9.7483
#2	9.8699
#3	9.6738
#4	9.9919
#5	9.9287

3116

Method: ISOU Sample Name: prepstd

Operator: bjg

Comment: 201 df

Run Time: 03/03/99 11:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0002	.0035	.3564	.0050	51.62
Stddev	.0006	.0005	.0080	.0008	.28
%RSD	353.4	15.53	2.250	14.96	.5459

#1	.0003	.0041	.3624	.0051	51.66
#2	-.0010	.0034	.3448	.0061	51.84
#3	-.0006	.0027	.3512	.0041	51.81
#4	.0002	.0034	.3604	.0045	51.68
#5	.0001	.0039	.3629	.0053	51.14

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.086
Stddev	.064
%RSD	.63082

#1	10.002
#2	10.086
#3	10.069
#4	10.092
#5	10.180

HNF-1663 REV. 0

Method: ISOU Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/03/99 11:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	-.0008	-.0006	-.0008	.0021
Stddev	.0008	.0006	.0006	.0003	.0006
%RSD	170.3	71.66	100.9	34.31	27.03

#1	-.0015	-.0012	-.0017	-.0012	.0024
#2	.0007	-.0015	-.0003	-.0006	.0023
#3	-.0003	-.0003	.0001	-.0009	.0024
#4	-.0003	-.0006	-.0006	-.0007	.0024
#5	-.0009	-.0003	-.0007	-.0006	.0011

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.960
Stddev	1.402
%RSD	12.796

#1	13.396
#2	10.011
#3	10.019
#4	10.697
#5	10.677

3118

Method: ISOU

Sample Name: s98t003399

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99

11:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0006	-.0002	.0644	.0005	10.07
Stddev	.0003	.0004	.0026	.0010	.40
%RSD	47.11	263.5	4.114	207.2	3.938
#1	-.0009	-.0001	.0614	-.0001	9.445
#2	-.0008	.0005	.0681	-.0003	10.27
#3	-.0008	-.0004	.0656	.0004	9.954
#4	-.0002	-.0004	.0624	.0021	10.47
#5	-.0006	-.0004	.0643	.0001	10.22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.896
Stddev	.480
%RSD	4.4031

#1	11.738
#2	10.826
#3	10.713
#4	10.606
#5	10.598

Method: ISOU Sample Name: s98t003399_d Operator: bjg
Comment: 20301 df
Run Time: 03/03/99 11:21 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0004	.0776	.0012	12.30
Stddev	.0006	.0004	.0028	.0003	.26
%RSD	149.9	97.64	3.551	21.38	2.144
#1	-.0002	-.0007	.0811	.0009	12.65
#2	-.0004	.0000	.0743	.0011	12.41
#3	-.0007	-.0009	.0770	.0014	11.95
#4	-.0012	-.0001	.0761	.0012	12.16
#5	.0005	-.0002	.0797	.0015	12.34

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.415
Stddev	.147
%RSD	1.4160

#1	10.190
#2	10.524
#3	10.569
#4	10.389
#5	10.402

Method: ISOU Sample Name: s98t003399_a Operator: bjg
Comment: 20301 df
Run Time: 03/03/99 11:22 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	.0013	.1992	.0036	30.19
Stddev	.0005	.0006	.0086	.0006	1.24
%RSD	194.2	42.31	4.310	17.53	4.092

#1	-.0009	.0012	.1840	.0029	28.07
#2	-.0007	.0023	.2011	.0033	30.60
#3	.0000	.0009	.2040	.0036	31.14
#4	.0004	.0014	.2030	.0034	30.91
#5	-.0001	.0009	.2039	.0046	30.22

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.726
Stddev	.493
%RSD	4.6005

#1	11.601
#2	10.508
#3	10.407
#4	10.564
#5	10.547

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t003399_s

Operator: bjg

Comment: 20301 df

Run Time: 03/03/99 11:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	.0005	.0818	.0014	12.73
Stddev	.0005	.0006	.0015	.0007	.22
%RSD	134.5	108.2	1.850	47.70	1.698
#1	.0006	.0000	.0840	.0012	12.57
#2	-.0007	.0004	.0826	.0012	12.59
#3	-.0008	.0003	.0816	.0012	12.96
#4	-.0006	.0005	.0808	.0007	12.55
#5	-.0005	.0015	.0802	.0025	12.96

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.360
Stddev	.051
%RSD	.49355

#1	10.430
#2	10.376
#3	10.290
#4	10.362
#5	10.341

3122

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002930

Operator: bjg

Comment: 2010 df

Run Time: 03/03/99 11:32 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0023	.3039	.0081	46.99
Stddev	.0008	.0006	.0036	.0006	.44
%RSD	1880.	28.64	1.184	7.227	.9408

#1	.0001	.0021	.3014	.0071	46.39
#2	.0010	.0031	.3038	.0083	47.57
#3	-.0012	.0016	.2999	.0086	47.24
#4	.0000	.0028	.3092	.0080	46.90
#5	.0004	.0019	.3052	.0086	46.88

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.328
Stddev	.072
%RSD	.69832

#1	10.425
#2	10.225
#3	10.307
#4	10.338
#5	10.344

HNF-1663 REV. 0

Method: ISOU Sample Name: s9st002930_d Operator: bjg
Comment: 2010 df
Run Time: 03/03/99 11:35 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0003	.0020	.1869	.0048	29.77
Stddev	.0004	.0005	.0070	.0009	.59
%RSD	143.7	24.89	3.772	17.93	1.977
#1	.0003	.0019	.1803	.0041	30.46
#2	-.0001	.0023	.1820	.0052	29.03
#3	-.0008	.0012	.1941	.0041	29.29
#4	-.0005	.0021	.1949	.0044	29.99
#5	-.0004	.0024	.1831	.0060	30.08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.467
Stddev	.097
%RSD	.92910

#1	10.335
#2	10.553
#3	10.536
#4	10.391
#5	10.518

3124

Method: ISOU Sample Name: s98t002930_a Operator: bjg
Comment: 2010 df
Run Time: 03/03/99 11:38 Type: Unk Mode: CONC Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0000	.0032	.4416	.0104	67.88
Stddev	.001	.0006	.0066	.0009	.38
%RSD	3253.	18.64	1.484	8.802	.5643

#1	.0002	.0039	.4394	.0110	67.36
#2	.0000	.0024	.4461	.0091	68.15
#3	.0000	.0033	.4338	.0104	68.21
#4	-.0008	.0029	.4385	.0114	67.58
#5	.0006	.0036	.4503	.0099	68.10

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.123
Stddev	.043
%RSD	.42723

#1	10.153
#2	10.094
#3	10.111
#4	10.075
#5	10.181

Method: ISOU   Sample Name: s98t002930_s   Operator: bjg
Comment: 2010 df
Run Time: 03/03/99 11:41 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0001	.0015	.2269	.0052	34.70
Stddev	.0003	.0002	.0045	.0008	.33
%RSD	241.2	16.96	1.979	16.08	.9581

#1	.0001	.0011	.2306	.0057	34.82
#2	.0004	.0016	.2316	.0037	34.55
#3	.0005	.0018	.2214	.0055	34.24
#4	-.0001	.0013	.2232	.0058	34.75
#5	-.0002	.0015	.2278	.0055	35.14

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.220
Stddev	.066
%RSD	.64657

#1	10.243
#2	10.153
#3	10.207
#4	10.322
#5	10.176

Method: ISOU Sample Name: ccv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/03/99 11:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0007	.0010	.1524	.0015	22.53
Stddev	.0001	.0007	.0029	.0007	.36
%RSD	19.84	69.08	1.913	48.07	1.577

#1	-.0007	.0017	.1491	.0018	22.23
#2	-.0009	.0006	.1542	.0010	22.41
#3	-.0006	.0019	.1554	.0026	22.51
#4	-.0008	.0007	.1538	.0009	22.37
#5	-.0005	.0003	.1494	.0012	23.14

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.9692
Stddev	.0602
%RSD	.60333

#1	9.9185
#2	10.058
#3	9.9105
#4	9.9663
#5	9.9928

HNF-1663 REV.0

Method: ISOU Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/03/99 11:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0009	-.0006	-.0005	-.0010	-.0001
Stddev	.0002	.0002	.0007	.0004	.0003
%RSD	19.78	26.58	137.2	38.94	214.0

#1	-.0008	-.0008	-.0009	-.0006	-.0002
#2	-.0011	-.0006	.0003	-.0009	.0001
#3	-.0007	-.0005	-.0014	-.0011	-.0001
#4	-.0011	-.0004	-.0007	-.0015	-.0006
#5	-.0011	-.0008	.0002	-.0007	.0001

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	10.280
Stddev	.100
%RSD	.96980

#1	10.413
#2	10.274
#3	10.247
#4	10.323
#5	10.142

LABCORE Completed Worklist Report for Worklist# 28448

Analyst: bjg

Instrument: ICPMS1

Book#: 8B59, 990303-3
11B59 (deleted 1:1)

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSAC-QC TH232	QC	0.50e-2	5.35e-03	107.000 % Recovery	
1	ICV	0	EMSAC-QC NP237	QC	4.94e-3	5.96e-03	120.648 % Recovery	
1	ICV	0	EMSAC-QC PU239	QC	0.50e-2	5.11e-03	102.200 % Recovery	
1	ICV	0	EMSAC-QC 241	QC	0.513e-2	5.56e-03	108.382 % Recovery	
2	ICB	0	EMSAC-QC PA231	QC	1	<2.29e-6	ug/mL	
2	ICB	0	EMSAC-QC TH229	QC	1	<3.27e-6	ug/mL	
2	ICB	0	EMSAC-QC TH230	QC	1	<3.67e-6	ug/mL	
2	ICB	0	EMSAC-QC TH232	QC	1	<3.44e-6	ug/mL	
2	ICB	0	EMSAC-QC NP237	QC	1	<5.74e-6	ug/mL	
2	ICB	0	EMSAC-QC PU239	QC	1	<6.51e-6	ug/mL	
2	ICB	0	EMSAC-QC PU240	QC	1	<3.88e-6	ug/mL	
2	ICB	0	EMSAC-QC PU242	QC	1	<3.80e-6	ug/mL	
2	ICB	0	EMSAC-QC 241	QC	1	<3.66e-6	ug/mL	
2	ICB	0	EMSAC-QC 243	QC	1	<3.26e-6	ug/mL	
2	ICB	0	EMSAC-QC 244	QC	1	<2.71e-6	ug/mL	
3	SAMPLE	S98T002965	0 A EMSACTAL PA231-A1	LIQUID	N/A	<9.262e-02	9.26e-002	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL TH229-A1	LIQUID	N/A	<1.325e-01	0.133	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL TH230-A1	LIQUID	N/A	<1.486e-01	0.149	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL TH232-A1	LIQUID	N/A	<1.394e-01	0.139	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL NP237-A1	LIQUID	N/A	<2.324e-01	0.232	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL PU239-A1	LIQUID	N/A	<2.638e-01	0.264	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL PU240-A1	LIQUID	N/A	<1.572e-01	0.157	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL PU242-A1	LIQUID	N/A	<1.538e-01	0.154	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL 241-A1	LIQUID	N/A	<1.484e-01	0.148	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL 243-A1	LIQUID	N/A	<1.318e-01	0.132	ug/mL
3	SAMPLE	S98T002965	0 A EMSACTAL 244-A1	LIQUID	N/A	<1.097e-01	0.110	ug/mL
4	DUP	S98T002965	0 A EMSACTAL PA231-A1	LIQUID	<9.26e-2	<9.26e-2	RPD	
4	DUP	S98T002965	0 A EMSACTAL TH229-A1	LIQUID	<1.33e-1	<1.32e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL TH230-A1	LIQUID	<1.49e-1	<1.49e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL TH232-A1	LIQUID	<1.39e-1	<1.39e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL NP237-A1	LIQUID	<2.32e-1	<2.32e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL PU239-A1	LIQUID	<2.64e-1	<2.64e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL PU240-A1	LIQUID	<1.57e-1	<1.57e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL PU242-A1	LIQUID	<1.54e-1	<1.54e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL 241-A1	LIQUID	<1.48e-1	<1.48e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL 243-A1	LIQUID	<1.32e-1	<1.32e-1	RPD	
4	DUP	S98T002965	0 A EMSACTAL 244-A1	LIQUID	<1.10e-1	<1.10e-1	RPD	
5	SPK-POST	S98T002965	0 A EMSACTAL TH232-A1	LIQUID	0.0200	1.88e-02	94.000 % Recovery	
5	SPK-POST	S98T002965	0 A EMSACTAL NP237-A1	LIQUID	0.01976	2.16e-02	109.312 % Recovery	
5	SPK-POST	S98T002965	0 A EMSACTAL PU239-A1	LIQUID	0.02008	1.89e-02	94.124 % Recovery	
5	SPK-POST	S98T002965	0 A EMSACTAL 241-A1	LIQUID	0.02052	2.05e-02	99.903 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28448

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 SAMPLE	S98T002966	0 A	EMSACTAL PA231-A1	LIQUID	N/A <	9.262e-02	9.26e-002	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL TH229-A1	LIQUID	N/A <	1.325e-01	0.133	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL TH230-A1	LIQUID	N/A <	1.486e-01	0.149	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL TH232-A1	LIQUID	N/A <	1.394e-01	0.139	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL NP237-A1	LIQUID	N/A <	2.324e-01	0.232	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL FU239-A1	LIQUID	N/A <	2.638e-01	0.264	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL PU240-A1	LIQUID	N/A <	1.572e-01	0.157	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL PU242-A1	LIQUID	N/A <	1.538e-01	0.154	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL 241-A1	LIQUID	N/A <	1.484e-01	0.148	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL 243-A1	LIQUID	N/A <	1.318e-01	0.132	ug/mL
6 SAMPLE	S98T002966	0 A	EMSACTAL 244-A1	LIQUID	N/A <	1.097e-01	0.110	ug/mL
7 DUP	S98T002966	0 A	EMSACTAL PA231-A1	LIQUID	<9.26e-2	<9.26e-2		RPD
7 DUP	S98T002966	0 A	EMSACTAL TH229-A1	LIQUID	<1.33e-1	<1.32e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL TH230-A1	LIQUID	<1.49e-1	<1.49e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL TH232-A1	LIQUID	<1.39e-1	<1.39e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL NP237-A1	LIQUID	<2.32e-1	<2.32e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL FU239-A1	LIQUID	<2.64e-1	<2.64e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL PU240-A1	LIQUID	<1.57e-1	<1.57e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL PU242-A1	LIQUID	<1.54e-1	<1.54e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL 241-A1	LIQUID	<1.48e-1	<1.48e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL 243-A1	LIQUID	<1.32e-1	<1.32e-1		RPD
7 DUP	S98T002966	0 A	EMSACTAL 244-A1	LIQUID	<1.10e-1	<1.10e-1		RPD
8 SAMPLE	S98T002967	0 A	EMSACTAL PA231-A1	LIQUID	N/A <	9.262e-02	9.26e-002	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL TH229-A1	LIQUID	N/A <	1.325e-01	0.133	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL TH230-A1	LIQUID	N/A <	1.486e-01	0.149	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL TH232-A1	LIQUID	N/A <	1.394e-01	0.139	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL NP237-A1	LIQUID	N/A <	2.324e-01	0.232	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL FU239-A1	LIQUID	N/A <	2.638e-01	0.264	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL PU240-A1	LIQUID	N/A <	1.572e-01	0.157	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL PU242-A1	LIQUID	N/A <	1.538e-01	0.154	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL 241-A1	LIQUID	N/A <	1.484e-01	0.148	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL 243-A1	LIQUID	N/A <	1.318e-01	0.132	ug/mL
8 SAMPLE	S98T002967	0 A	EMSACTAL 244-A1	LIQUID	N/A <	1.097e-01	0.110	ug/mL
9 DUP	S98T002967	0 A	EMSACTAL PA231-A1	LIQUID	<9.26e-2	<9.26e-2		RPD
9 DUP	S98T002967	0 A	EMSACTAL TH229-A1	LIQUID	<1.33e-1	<1.32e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL TH230-A1	LIQUID	<1.49e-1	<1.49e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL TH232-A1	LIQUID	<1.39e-1	<1.39e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL NP237-A1	LIQUID	<2.32e-1	<2.32e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL FU239-A1	LIQUID	<2.64e-1	<2.64e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL PU240-A1	LIQUID	<1.57e-1	<1.57e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL PU242-A1	LIQUID	<1.54e-1	<1.54e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL 241-A1	LIQUID	<1.48e-1	<1.48e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL 243-A1	LIQUID	<1.32e-1	<1.32e-1		RPD
9 DUP	S98T002967	0 A	EMSACTAL 244-A1	LIQUID	<1.10e-1	<1.10e-1		RPD
10 SPK-POST	S98T002967	0 A	EMSACTAL TH232-A1	LIQUID	0.020	1.97e-02	98.500 % Recovery	
10 SPK-POST	S98T002967	0 A	EMSACTAL NP237-A1	LIQUID	0.01976	2.27e-02	114.879 % Recovery	
10 SPK-POST	S98T002967	0 A	EMSACTAL FU239-A1	LIQUID	0.02008	1.96e-02	97.610 % Recovery	
10 SPK-POST	S98T002967	0 A	EMSACTAL 241-A1	LIQUID	0.02052	2.11e-02	102.827 % Recovery	
11 CVV		0	EMSAC-QC TH232	QC	0.50e-2	5.33e-03	106.600 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

3130

LABCORE Completed Worklist Report for Worklist# 28448

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	CCV	0	EMSAC-QC NP237	QC	4.940e-3	6.07e-03	122.874 % Recovery	
11	CCV	0	EMSAC-QC PU239	QC	0.500e-2	5.31e-03	106.200 % Recovery	
11	CCV	0	EMSAC-QC 241	QC	0.5130e-2	5.69e-03	110.916 % Recovery	
12	CCB	0	EMSAC-QC PA231	QC	1	<2.29e-6		ug/mL
12	CCB	0	EMSAC-QC TH229	QC	1	<3.27e-6		ug/mL
12	CCB	0	EMSAC-QC TH230	QC	1	<3.67e-6		ug/mL
12	CCB	0	EMSAC-QC TH232	QC	1	<3.44e-6		ug/mL
12	CCB	0	EMSAC-QC NP237	QC	1	<5.74e-6		ug/mL
12	CCB	0	EMSAC-QC PU239	QC	1	<6.51e-6		ug/mL
12	CCB	0	EMSAC-QC PU240	QC	1	<3.88e-6		ug/mL
12	CCB	0	EMSAC-QC PU242	QC	1	<3.80e-6		ug/mL
12	CCB	0	EMSAC-QC 241	QC	1	<3.66e-6		ug/mL
12	CCB	0	EMSAC-QC 243	QC	1	<3.26e-6		ug/mL
12	CCB	0	EMSAC-QC 244	QC	1	<2.71e-6		ug/mL

Final page for worklist# 28448

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul M. Pang 03/04/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28448

Analyst: B56 Instrument: ICPMS1 Book# 8857 SS0303-3 (11857)
 Method: LA-506-101 Rev/Mod A-0 DLUTEA 1:1

Worklist Comment: ICP/MS BY-105 (DIRECT) ACTINIDE

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSAC-QC	QC		
2 ICB			@MSAC-QC	QC		
3 SAMPLE	S98T002965 0 A		@MSACTA1 LIQUID		98000509 BY-105 (2)	
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
4 DUP	S98T002965 0 A		@MSACTA1 LIQUID			
5 SPK-POST	S98T002965 0 A		@MSACTA1 LIQUID			
6 SAMPLE	S98T002966 0 A		@MSACTA1 LIQUID		98000509 BY-105 (2)	
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
7 DUP	S98T002966 0 A		@MSACTA1 LIQUID			
8 SAMPLE	S98T002967 0 A		@MSACTA1 LIQUID		98000509 BY-105 (2)	
Analytes Requested: 241-A1 , 243-A1 , 244-A1 , NP237-A1, PA231-A1, PU239-A1, PU240-A1, PU242-A1, TH229-A1, TH230-A1, TH232-A1						
9 DUP	S98T002967 0 A		@MSACTA1 LIQUID			
10 SPK-POST	S98T002967 0 A		@MSACTA1 LIQUID			
11 CCV			@MSAC-QC	QC		
12 CCB			@MSAC-QC	QC		

Final page for worklist # 28448

[Signature] 3-4-99
 Signature Date

 Signature Date

Data Entry Comments:

SS8T002965-67

1-10-1-1025-10-1

POST spikes = 04-1 of 8857 is 100.

DE
40501

Uploaded & Validated by:
Sam M. Lamy
 03/04/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3132

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/04/99 09:57 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	155700.
Stddev	1158.
%RSD	.7434

#1	155300.
#2	154200.
#3	155700.
#4	157700.
#5	156200.
#6	154700.
#7	156900.
#8	156800.
#9	154300.
#10	155500.

3-4-99

By - 105

ACTIVE

598T00 2865-67

FILE # AC0304A.TXT

1158 MSRT. CTS PLUTED 1:1

WL# 28448

990303-3

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3133 TO 3165.

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: s98t002967_a

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:20 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2661	.4198	1679.	4.943	2.436
Stddev	.0356	.0191	56.	.126	.099
%RSD	13.38	4.539	3.348	2.545	4.067

#1	.2669	.3960	1586.	4.750	2.307
#2	.2904	.4178	1682.	5.043	2.504
#3	.2045	.4327	1711.	4.946	2.359
#4	.2871	.4084	1683.	4.913	2.542
#5	.2817	.4441	1733.	5.065	2.465

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2049.	1801.	49.35	.4882	.2471
Stddev	40.	82.	1.58	.0391	.0052
%RSD	1.946	4.533	3.204	8.010	2.112

#1	1983.	1656.	46.72	.5338	.2411
#2	2042.	1826.	49.64	.4800	.2519
#3	2058.	1830.	49.31	.4564	.2489
#4	2081.	1839.	50.20	.5238	.2516
#5	2079.	1853.	50.86	.4471	.2419

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1981.	.2926
Stddev	43.	.0297
%RSD	2.157	10.15

#1	1916.	.2812
#2	1991.	.2696
#3	1984.	.3260
#4	1976.	.3226
#5	2036.	.2634

Sample Name: s98t002967_a Run Time: 03/04/99 11:20

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	112.52
Stddev	2.58
%RSD	2.2963

#1	116.15
#2	112.50
#3	112.48
#4	112.63
#5	108.84

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/04/99 11:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2731	.2850	455.8	7.751	.8502
Stddev	.0234	.0362	4.3	.052	.0363
%RSD	8.582	12.69	.9412	.6776	4.275

#1	.3069	.2527	451.6	7.756	.8364
#2	.2470	.2775	453.2	7.692	.7960
#3	.2854	.3314	461.7	7.829	.8931
#4	.2594	.2503	459.0	7.716	.8638
#5	.2668	.3128	453.6	7.765	.8618

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	547.8	489.0	13.38	.3379	.2768
Stddev	4.8	7.7	.25	.0119	.0360
%RSD	.8760	1.578	1.857	3.508	13.00

#1	541.0	482.0	13.18	.3219	.2978
#2	548.1	489.9	13.75	.3446	.3111
#3	545.3	493.3	13.21	.3314	.2915
#4	553.0	480.8	13.23	.3388	.2625
#5	551.4	499.1	13.51	.3527	.2208

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	534.1	.2586
Stddev	3.8	.0234
%RSD	.7118	9.032

#1	529.1	.2497
#2	534.2	.2287
#3	533.3	.2578
#4	534.3	.2930
#5	539.8	.2637

Sample Name: ccv Run Time: 03/04/99 11:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.32
Stddev	.88
%RSD	.80328

#1	110.80
#2	109.30
#3	108.62
#4	109.21
#5	108.69

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/04/99 11:34 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2479	.2828	.4830	.3429	.2662
Stddev	.0232	.0295	.0329	.0373	.0568
%RSD	9.350	10.43	6.818	10.88	21.33

#1	.2202	.2516	.5253	.3995	.3177
#2	.2705	.2578	.4456	.3406	.2801
#3	.2701	.2987	.4989	.2955	.1907
#4	.2503	.3231	.4910	.3326	.2249
#5	.2287	.2826	.4541	.3462	.3176

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2561	.7598	.2740	.2639	.2921
Stddev	.0500	.1040	.0449	.0350	.0534
%RSD	19.53	13.69	16.37	13.25	18.29

#1	.2705	.9373	.1982	.2139	.3586
#2	.1719	.7193	.3024	.3087	.2419
#3	.2987	.7626	.3082	.2765	.2987
#4	.2534	.7033	.2915	.2503	.3263
#5	.2858	.6764	.2699	.2699	.2350

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2917	.2643
Stddev	.0156	.0362
%RSD	5.348	13.71

#1	.2768	.2831
#2	.2864	.2419
#3	.3178	.3082
#4	.2851	.2154
#5	.2922	.2731

Sample Name: ccb

Run Name: 03/04/99 11:34

~~HNF-1663 REV. 0~~~~HNF-1663 REV. 0~~

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	105.16
Stddev	.49
%RSD	.46659

#1	105.98
#2	104.73
#3	104.90
#4	105.22
#5	104.96

Operator: bjg

Corr.Fact: 1.000000

#1	155300.
#2	154200.
#3	155700.
#4	157700.
#5	156200.
#6	154700.
#7	156900.
#8	156800.
#9	154300.
#10	155500.

B. K. C.
3-4-88

3-4-88

By - 205

Activities

598700 2865-67

FILE # AC 0304 A 725

11B59 MSRT. LCS DATED 1:1
950303-3

Method: ACTINIDE Sample Name: cal blank

Operator: bjg

Comment: background

Run Time: 03/04/99 09:58 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3589	.3551	.4730	.4184	.3913
Stddev	.0386	.0438	.0420	.0409	.0276
%RSD	10.77	12.33	8.870	9.767	7.050

#1	.3478	.3739	.5130	.4000	.3696
#2	.3014	.3694	.4458	.3651	.4033
#3	.3757	.2818	.5209	.4739	.3714
#4	.3633	.3967	.4260	.4385	.4343
#5	.4063	.3535	.4591	.4144	.3778

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3916	.6000	.4456	.3847	.4157
Stddev	.0517	.0599	.0515	.0514	.0374
%RSD	13.21	9.987	11.55	13.37	8.996

#1	.4696	.5826	.4956	.4000	.4739
#2	.3524	.6623	.4797	.4161	.3863
#3	.3373	.5550	.3629	.4440	.4312
#4	.3967	.6640	.4385	.3425	.4009
#5	.4022	.5363	.4510	.3210	.3860

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3931	.3837
Stddev	.0491	.0445
%RSD	12.48	11.59

#1	.4130	.3783
#2	.4373	.4415
#3	.3117	.3501
#4	.4176	.3341
#5	.3860	.4144

Sample Name: cal blank Run Time: 03/04/99 09:58

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	79.024
Stddev	2.029
%RSD	2.5676

#1	76.668
#2	78.515
#3	78.075
#4	79.817
#5	82.045

Method: ACTINIDE Sample Name: cal std
 Comment: 0.010 ppm cal
 Run Time: 03/04/99 10:07 Type: Unk

Operator: bjg

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3269	.4093	854.6	910.6	1.679
Stddev	.0489	.0211	7.6	9.4	.051
%RSD	14.96	5.164	.8909	1.034	3.037
#1	.3896	.3857	845.2	900.3	1.713
#2	.3595	.4390	864.5	919.9	1.608
#3	.2801	.4221	859.8	920.7	1.658
#4	.3276	.3975	852.7	903.0	1.741
#5	.2780	.4023	850.8	908.9	1.675

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.995	.7126	.3530	.3202	.3299
Stddev	.118	.0248	.0603	.0128	.0396
%RSD	5.935	3.477	17.08	4.011	12.00
#1	2.145	.7406	.3086	.3009	.3124
#2	2.036	.7001	.3027	.3254	.3784
#3	2.006	.7246	.3212	.3361	.2764
#4	1.973	.7214	.3938	.3202	.3570
#5	1.818	.6766	.4389	.3182	.3255

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3508	.3410
Stddev	.0179	.0229
%RSD	5.116	6.715
#1	.3240	.3240
#2	.3633	.3141
#3	.3698	.3698
#4	.3533	.3570
#5	.3438	.3401

Sample Name: cal std Run Time: 03/04/99 10:07

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	89.092
Stdddev	1.908
%RSD	2.1419

#1	86.418
#2	88.087
#3	89.247
#4	90.570
#5	91.140

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 03/04/99 10:15 Type: Unk

Mode: IR

Corr. Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4873	1.365	8742.	9021.	14.09
Stddev	.0205	.046	173.	209.	.50
%RSD	4.211	3.397	1.979	2.319	3.539
#1	.5171	1.339	8944.	8920.	14.84
#2	.4976	1.383	8637.	9078.	14.16
#3	.4725	1.433	8918.	9330.	14.14
#4	.4833	1.362	8615.	9013.	13.78
#5	.4659	1.311	8598.	8764.	13.52

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	16.46	2.771	.3738	.3445	.3522
Stddev	.68	.103	.0418	.0210	.0390
%RSD	4.108	3.721	11.19	6.091	11.08

#1	16.90	2.751	.4399	.3203	.3666
#2	16.48	2.678	.3494	.3532	.3418
#3	17.24	2.882	.3898	.3741	.3544
#4	16.19	2.670	.3529	.3452	.4028
#5	15.49	2.875	.3371	.3296	.2955

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3289	.3511
Stddev	.0168	.0191
%RSD	5.123	5.437

#1	.3550	.3511
#2	.3267	.3760
#3	.3111	.3229
#4	.3184	.3567
#5	.3333	.3485

Sample Name: cal std Run Time: 03/04/99 10:15

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	86.740
Stddev	1.333
%RSD	1.5367

#1	86.385
#2	87.759
#3	84.658
#4	86.902
#5	87.998

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact lcs

Run Time: 03/04/99 10:26 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3123	.3663	457.6	7.788	.9341
Stddev	.0232	.0213	9.0	.266	.0600
%RSD	7.421	5.817	1.976	3.416	6.426

#1	.3020	.3476	465.2	7.813	.9797
#2	.2908	.3889	442.3	7.344	.9356
#3	.3080	.3570	462.5	8.036	.8750
#4	.3088	.3895	457.0	7.805	.8738
#5	.3517	.3482	461.1	7.942	1.006

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	537.2	471.1	13.06	.4187	.3256
Stddev	8.7	5.2	.32	.0579	.0209
%RSD	1.616	1.112	2.428	13.83	6.434

#1	545.4	463.6	12.96	.3476	.3301
#2	528.8	470.2	12.96	.4415	.3154
#3	539.0	475.1	12.66	.3675	.3465
#4	545.3	469.8	13.22	.4562	.2948
#5	527.5	477.0	13.51	.4805	.3412

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	522.0	.3264
Stddev	7.3	.0312
%RSD	1.399	9.568

#1	517.6	.3336
#2	527.3	.3539
#3	512.1	.3080
#4	523.0	.3544
#5	530.2	.2820

Sample Name: icv Run Time: 03/04/99 10:26

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	95.202
Stddev	.321
%RSD	.33671

#1	94.929
#2	95.128
#3	95.235
#4	94.983
#5	95.733

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/04/99 10:36 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3160	.3210	.5947	.3749	.3211
Stddev	.0544	.0188	.0436	.0111	.0327
%RSD	17.21	5.864	7.336	2.973	10.19

#1	.3065	.3306	.5821	.3823	.3341
#2	.2508	.3037	.6675	.3673	.3249
#3	.4006	.3031	.5991	.3832	.2996
#4	.3230	.3475	.5582	.3826	.2808
#5	.2991	.3202	.5665	.3589	.3659

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.3136	.4750	.3595	.3477	.3196
Stddev	.0482	.0296	.0539	.0435	.0260
%RSD	15.36	6.240	14.99	12.51	8.134

#1	.2928	.4684	.3306	.3754	.3031
#2	.3885	.4839	.3426	.3708	.3002
#3	.2787	.4598	.4250	.3031	.3518
#4	.2738	.4423	.2949	.2984	.3440
#5	.3343	.5208	.4046	.3906	.2991

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3477	.3512
Stddev	.0411	.0283
%RSD	11.82	8.049

#1	.3444	.3616
#2	.4097	.3673
#3	.3379	.3170
#4	.2949	.3265
#5	.3519	.3835

Sample Name: icb Run Time: 03/04/99 10:36

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	95.308
Stddev	.953
%RSD	.99986

#1	96.780
#2	94.379
#3	95.692
#4	94.955
#5	94.734

Method: ACTINIDE Sample Name: s98t002965

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 10:41 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2848	.2639	.4865	1.767	.2640
Stddev	.0430	.0390	.0235	.114	.0358
%RSD	15.11	14.79	4.837	6.450	13.58

#1	.2633	.3274	.5097	1.839	.2633
#2	.2395	.2428	.4692	1.906	.2034
#3	.2597	.2337	.5096	1.740	.2921
#4	.3276	.2757	.4865	1.605	.2887
#5	.3341	.2401	.4574	1.745	.2725

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2777	.3796	.2924	.2685	.2818
Stddev	.0185	.0210	.0266	.0409	.0525
%RSD	6.663	5.519	9.107	15.22	18.63

#1	.2498	.3848	.3038	.3240	.3679
#2	.2887	.4101	.2723	.2658	.2362
#3	.2986	.3830	.2564	.2597	.2759
#4	.2724	.3633	.3178	.2108	.2432
#5	.2790	.3569	.3114	.2822	.2855

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.3092	.3068
Stddev	.0507	.0209
%RSD	16.39	6.818

#1	.3848	.3173
#2	.2625	.2789
#3	.2629	.3149
#4	.3243	.3308
#5	.3114	.2920

Sample Name: s98t002965 Run Time: 03/04/99 10:41

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	101.71
Stddev	1.73
%RSD	1.6960

#1	98.759
#2	101.59
#3	102.69
#4	102.78
#5	102.75

Method: ACTINIDE Sample Name: s98t002965_d

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 10:45 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2702	.2791	.4489	1.763	.2535
Stddev	.0354	.0545	.0181	.054	.0143
%RSD	13.12	19.51	4.022	3.058	5.630

#1	.3301	.2916	.4486	1.721	.2532
#2	.2622	.3326	.4541	1.762	.2750
#3	.2427	.2491	.4184	1.808	.2555
#4	.2453	.2007	.4587	1.825	.2484
#5	.2705	.3215	.4647	1.700	.2355

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2726	.3512	.2746	.2752	.2937
Stddev	.0399	.0383	.0335	.0198	.0287
%RSD	14.64	10.90	12.18	7.209	9.785

#1	.2980	.4038	.2275	.2916	.2852
#2	.2047	.3230	.2942	.2718	.2846
#3	.2874	.3353	.2938	.2427	.3353
#4	.2707	.3153	.3058	.2898	.2580
#5	.3024	.3788	.2515	.2801	.3056

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2842	.2867
Stddev	.0280	.0190
%RSD	9.869	6.635

#1	.2916	.2948
#2	.2654	.2814
#3	.2874	.3130
#4	.3249	.2612
#5	.2515	.2833

Sample Name: s98t002965_d Run Time: 03/04/99 10:45

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	104.40
Stddev	.29
%RSD	.28104

#1	104.02
#2	104.24
#3	104.37
#4	104.65
#5	104.73

Method: ACTINIDE Sample Name: s98t002965_a

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 10:51 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2825	.4327	1609.	5.309	2.441
Stddev	.0446	.0255	17.	.108	.058
%RSD	15.78	5.904	1.036	2.043	2.358

#1	.3453	.4489	1581.	5.371	2.426
#2	.2244	.4582	1621.	5.233	2.503
#3	.2751	.3921	1612.	5.467	2.444
#4	.3008	.4370	1611.	5.203	2.353
#5	.2671	.4273	1622.	5.273	2.479

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1950.	1740.	47.16	.5141	.2687
Stddev	23.	19.	.09	.0486	.0119
%RSD	1.187	1.118	.1978	9.457	4.436

#1	1915.	1724.	47.14	.5587	.2511
#2	1966.	1732.	47.02	.4677	.2781
#3	1943.	1736.	47.20	.5502	.2814
#4	1975.	1773.	47.26	.5384	.2660
#5	1953.	1733.	47.20	.4556	.2671

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1923.	.2870
Stddev	18.	.0244
%RSD	.9598	8.491

#1	1900.	.2982
#2	1950.	.2591
#3	1917.	.3225
#4	1931.	.2818
#5	1919.	.2734

Sample Name: s98t002965_a Run Time: 03/04/99 10:51

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	105.69
Stddev	.42
%RSD	.39748

#1	106.19
#2	105.48
#3	105.42
#4	105.26
#5	106.08

Method: ACTINIDE Sample Name: s98t002966

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:03 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2815	.2618	.4217	1.556	.2486
Stddev	.0317	.0206	.0313	.084	.0447
%RSD	11.25	7.884	7.427	5.370	18.00

#1	.3097	.2508	.4366	1.554	.3189
#2	.3050	.2628	.4681	1.435	.2658
#3	.2300	.2451	.3874	1.671	.2270
#4	.2840	.2535	.4092	1.558	.2229
#5	.2786	.2970	.4072	1.564	.2082

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2613	.6308	.2783	.2740	.2643
Stddev	.0169	.0565	.0239	.0151	.0313
%RSD	6.470	8.949	8.601	5.505	11.84

#1	.2880	.7215	.2601	.2601	.2632
#2	.2416	.5799	.2537	.2809	.2235
#3	.2633	.6083	.2784	.2572	.2814
#4	.2565	.6474	.3146	.2932	.3054
#5	.2572	.5970	.2847	.2786	.2480

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2709	.2833
Stddev	.0450	.0317
%RSD	16.61	11.18

#1	.2075	.3066
#2	.2658	.2748
#3	.2603	.2936
#4	.2901	.2321
#5	.3306	.3092

Sample Name: s98t002966 Run Time: 03/04/99 11:03

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.24
Stddev	1.09
%RSD	.99970

#1	107.65
#2	110.37
#3	110.14
#4	109.15
#5	108.88

Method: ACTINIDE Sample Name: s98t002966_d

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:07 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2794	.2704	.4013	1.574	.2523
Stddev	.0189	.0115	.0309	.073	.0408
%RSD	6.761	4.238	7.703	4.616	16.19

#1	.2614	.2857	.4012	1.471	.1945
#2	.2963	.2631	.3749	1.642	.2510
#3	.2572	.2751	.3917	1.630	.3020
#4	.2863	.2559	.4539	1.599	.2772
#5	.2960	.2723	.3848	1.527	.2368

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2909	.5638	.2764	.2422	.2573
Stddev	.0289	.0080	.0370	.0422	.0164
%RSD	9.920	1.428	13.38	17.41	6.385

#1	.2462	.5684	.2675	.2097	.2492
#2	.3054	.5503	.2963	.2359	.2812
#3	.3080	.5622	.3289	.2063	.2572
#4	.3168	.5696	.2345	.3107	.2620
#5	.2782	.5683	.2545	.2486	.2368

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2613	.2450
Stddev	.0398	.0306
%RSD	15.24	12.50

#1	.2310	.2462
#2	.2903	.1935
#3	.3170	.2721
#4	.2315	.2498
#5	.2368	.2634

HNF-1663 REV. 0

Sample Name: s98t002966_d Run Time: 03/04/99 11:07

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	110.69
Stddev	1.34
%RSD	1.2103

#1	109.67
#2	110.24
#3	111.47
#4	109.43
#5	112.62

Method: ACTINIDE Sample Name: s98t002967

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:12 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2609	.2343	.3795	1.556	.2384
Stddev	.0277	.0276	.0250	.022	.0349
%RSD	10.62	11.79	6.598	1.430	14.63

#1	.2339	.2308	.3478	1.557	.2185
#2	.2573	.2757	.3768	1.572	.1899
#3	.2746	.2417	.4148	1.519	.2537
#4	.2378	.2225	.3902	1.558	.2804
#5	.3010	.2007	.3679	1.575	.2493

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2690	.4814	.2762	.2461	.2624
Stddev	.0339	.0311	.0194	.0295	.0229
%RSD	12.58	6.451	7.020	11.98	8.730

#1	.2339	.4801	.2555	.2647	.2832
#2	.3217	.5147	.2635	.2727	.2910
#3	.2507	.4656	.2686	.1970	.2507
#4	.2804	.4390	.2957	.2469	.2439
#5	.2585	.5078	.2980	.2493	.2432

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2677	.2671
Stddev	.0363	.0256
%RSD	13.57	9.591

#1	.2493	.2832
#2	.2910	.2298
#3	.2656	.2716
#4	.2195	.2957
#5	.3132	.2554

Sample Name: s98t002967 Run Time: 03/04/99 11:12

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	109.56
Stddev	1.30
%RSD	1.1871

#1	108.30
#2	108.80
#3	111.70
#4	109.35
#5	109.63

HNF-1663 REV. 0

Method: ACTINIDE

Sample Name: s98t002967_d

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:16 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2452	.2710	.3417	1.542	.2607
Stddev	.0143	.0508	.0458	.092	.0357
%RSD	5.849	18.75	13.41	5.983	13.71

#1	.2595	.3222	.3461	1.456	.2894
#2	.2430	.2520	.3000	1.644	.2790
#3	.2374	.1953	.4176	1.544	.2614
#4	.2265	.3111	.3141	1.622	.1993
#5	.2594	.2743	.3309	1.443	.2743

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2332	.4982	.2542	.2771	.2524
Stddev	.0084	.0225	.0216	.0578	.0205
%RSD	3.608	4.519	8.495	20.87	8.138

#1	.2237	.5161	.2566	.2685	.2715
#2	.2400	.4800	.2790	.3240	.2730
#3	.2434	.4687	.2644	.3365	.2374
#4	.2325	.5074	.2507	.2658	.2265
#5	.2266	.5187	.2206	.1908	.2534

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.2668	.2680
Stddev	.0137	.0353
%RSD	5.131	13.18

#1	.2774	.2566
#2	.2580	.2760
#3	.2794	.2163
#4	.2718	.2778
#5	.2474	.3130

HNF-1663 REV.0

Sample Name: s98t002967_d Run Time: 03/04/99 11:16

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	111.19
Stddev	.59
%RSD	.53506

#1	111.73
#2	111.11
#3	110.95
#4	110.37
#5	111.81

Method: ACTINIDE Sample Name: s98t002967_a

Operator: bjg

Comment: 40501 df

Run Time: 03/04/99 11:20 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.2661	.4198	1679.	4.943	2.436
Stdddev	.0356	.0191	56.	.126	.099
%RSD	13.38	4.539	3.348	2.545	4.067
#1	.2669	.3960	1586.	4.750	2.307
#2	.2904	.4178	1682.	5.043	2.504
#3	.2045	.4327	1711.	4.946	2.359
#4	.2871	.4084	1683.	4.913	2.542
#5	.2817	.4441	1733.	5.065	2.465

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2049.	1801.	49.35	.4882	.2471
Stdddev	40.	82.	1.58	.0391	.0052
%RSD	1.946	4.533	3.204	8.010	2.112
#1	1983.	1656.	46.72	.5338	.2411
#2	2042.	1826.	49.64	.4800	.2519
#3	2058.	1830.	49.31	.4564	.2489
#4	2081.	1839.	50.20	.5238	.2516
#5	2079.	1853.	50.86	.4471	.2419

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1981.	.2926
Stdddev	43.	.0297
%RSD	2.157	10.15
#1	1916.	.2812
#2	1991.	.2696
#3	1984.	.3260
#4	1976.	.3226
#5	2036.	.2634

LABCORE Completed Worklist Report for Worklist# 28459

Analyst: bjg

Instrument: ICPMS1

Book#: 990301-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) Field Blank ISO-U

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	QMSU-QC U235	QC	0.000142	1.46e-04	102.817 % Recovery	
1	ICV	0	QMSU-QC U238	QC	0.020	2.05e-02	102.500 % Recovery	
2	ICB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
2	ICB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
2	ICB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	
3	SAMPLE	S98T002391 0 D	QMSU-D1 U233-D1	LIQUID	N/A	< 1.212e-03	1.21e-003	ug/mL
3	SAMPLE	S98T002391 0 D	QMSU-D1 U234-D1	LIQUID	N/A	< 1.212e-03	1.21e-003	ug/mL
3	SAMPLE	S98T002391 0 D	QMSU-D1 U235-D1	LIQUID	N/A	< 1.212e-03	1.21e-003	ug/mL
3	SAMPLE	S98T002391 0 D	QMSU-D1 U236-D1	LIQUID	N/A	< 1.616e-03	1.62e-003	ug/mL
3	SAMPLE	S98T002391 0 D	QMSU-D1 U238-D1	LIQUID	N/A	< 1.212e-03	1.21e-003	ug/mL
4	DUP	S98T002391 0 D	QMSU-D1 U233-D1	LIQUID	<1.21e-3	<1.21e-3	RPD	
4	DUP	S98T002391 0 D	QMSU-D1 U234-D1	LIQUID	<1.21e-3	<1.21e-3	RPD	
4	DUP	S98T002391 0 D	QMSU-D1 U235-D1	LIQUID	<1.21e-3	<1.21e-3	RPD	
4	DUP	S98T002391 0 D	QMSU-D1 U236-D1	LIQUID	<1.62e-3	<1.62e-3	RPD	
4	DUP	S98T002391 0 D	QMSU-D1 U238-D1	LIQUID	<1.21e-3	<1.21e-3	RPD	
5	CCV	0	QMSU-QC U235	QC	0.000142	1.47e-04	103.521 % Recovery	
5	CCV	0	QMSU-QC U238	QC	0.020	2.13e-02	106.500 % Recovery	
6	CCB	0	QMSU-QC U233	QC	1	<1.20e-5	ug/mL	
6	CCB	0	QMSU-QC U234	QC	1	<1.20e-5	ug/mL	
6	CCB	0	QMSU-QC U235	QC	1	<1.20e-5	ug/mL	
6	CCB	0	QMSU-QC U236	QC	1	<1.60e-5	ug/mL	
6	CCB	0	QMSU-QC U238	QC	1	<1.20e-5	ug/mL	

Final page for worklist# 28459

Analyst Signature Date

Analyst Signature Date

Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28459

Analyst: BIG Instrument: ICPMS1 Book# 990301-1

Method: LA-506-101 Rev/Mod A0

Worklist Comment: ICP/MS BY-105 (DIRECT) Field Blank ISO-U

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			@MSU-QC	QC		
2 ICB			@MSU-QC	QC		
3 SAMPLE	S98T002391	0 D	@MSU-D1	LIQUID	98000436	BY-105 (2)
Analytes Requested: U233-D1 , U234-D1 , U235-D1 , U236-D1 , U238-D1						
4 DUP	S98T002391	0 D	@MSU-D1	LIQUID		
5 CCV			@MSU-QC	QC		
6 CCB			@MSU-QC	QC		

Final page for worklist # 28459

Signature 3-8-99

Signature Date

S98T002391
2391-D

1-10-1

DF
101
101

Data Entry Comments:

Uploaded & Validated by:

Sal M. Pang

03/08/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/08/99 13:32 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	171200.
Stddev	1609.
%RSD	.9398

#1	171700.
#2	171400.
#3	171600.
#4	171000.
#5	167900.
#6	173000.
#7	172500.
#8	170700.
#9	172800.
#10	169100.

3-8-99

BY-101 Field Blank

SSSTW 2381

FILE# U0208A.TRT

W.C.H. 28459

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3168 TO 3181.

Method: ISOU

Sample Name: BLANK

Operator:

Comment:

Run Time: 03/08/99 13:33 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	4.419	4.021	4.300	4.296	5.029
Stddev	.601	.513	.520	.498	.581
%RSD	13.59	12.75	12.10	11.60	11.56
#1	3.870	3.989	4.942	4.584	5.537
#2	4.378	4.378	3.709	4.561	4.257
#3	4.098	4.675	4.733	3.463	5.368
#4	4.317	3.442	3.967	4.667	5.426
#5	5.432	3.621	4.147	4.205	4.556

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.4828
Stddev	.1740
%RSD	2.0509

#1	8.3980
#2	8.2225
#3	8.6626
#4	8.5705
#5	8.5607

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	912.8732	4.4191	1.0000000	03/08/99 13:35:05

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.4191
100ppb U	0.711	0.711	0	0	653.47

U [233]
0
(S) IR²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	913.4328	4.0212	1.0000000	03/08/99 13:35:05

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.0212
100ppb U	0.711	0.711	0	0	653.47

U [234]
0
(S) IR²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	913.0412	4.2996	1.0000000	03/08/99 13:35:05

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.2996
100ppb U	0.711	0.711	0	0	653.47

U [235]
0
(S) IR²⁰⁰¹⁰

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	913.0461	4.2961	1.0000000	03/08/99 13:35:05

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S) IR
Blank	0	0	0	0	4.2961
100ppb U	0.711	0.711	0	0	653.47

U [236]
0
(S) IR²⁰⁰¹⁰

Concentration

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [238]	811.0237	5.0286	1.0000000	03/08/99 13:35:05

Standard Name	Concentration		Difference		Signal
	Stated	Found	Conc	%	(S)IR
Blank	0	0	0	0	5.0286
100ppb U	99.3	99.3	0	0	80540

U [238]
 0
 886000000
 (S)IR Concentration

HNF-1663 REV. 0

HNF-1663 REV. 0

Method: ISOU

Sample Name: 100ppb U

Operator:

Comment:

Run Time: 03/08/99 13:36 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	691.4	691.4	691.4	691.4	87560.
Stddev	37.2	37.2	37.2	37.2	1522.
%RSD	5.379	5.379	5.379	5.379	1.738
#1	735.0	735.0	735.0	735.0	89300.
#2	669.9	669.9	669.9	669.9	86570.
#3	640.1	640.1	640.1	640.1	85620.
#4	698.5	698.5	698.5	698.5	88770.
#5	713.5	713.5	713.5	713.5	87560.

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	8.7417
Stddev	.1295
%RSD	1.4815

#1	8.5236
#2	8.7697
#3	8.8667
#4	8.7964
#5	8.7521

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [233]	966.1786	4.4191	1.0000000	03/08/99 13:37:19

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.4191
100ppb U	0.711	0.711	0	0	691.37

U [233]

(S) IR⁰₇₇₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [234]	966.7382	4.0212	1.0000000	03/08/99 13:37:19

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.0212
100ppb U	0.711	0.711	0	0	691.37

U [234]

(S) IR⁰₇₇₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [235]	966.3466	4.2996	1.0000000	03/08/99 13:37:19

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.2996
100ppb U	0.711	0.711	0	0	691.37

U [235]

(S) IR⁰₇₇₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_ [236]	966.3515	4.2961	1.0000000	03/08/99 13:37:19

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S) IR
Blank	0	0	0	0	4.2961
100ppb U	0.711	0.711	0	0	691.37

U [236]

(S) IR⁰₇₇₀₀₁₀

Concentration

El Name	Slope	Y-int	Correlation	Date Stdized
U_[238]	881.7623	5.0286	1.0000000	03/08/99 13:37:19

Standard Name	Concentration Stated	Concentration Found	Difference Conc	Difference %	Signal (S)IR
Blank	0	0	0	0	5.0286
100ppb U	99.3	99.3	0	0	87564

U [238]

0

96320110
(S)IR

Concentration

Method: ISOU Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/08/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0004	-.0003	-.0009	.0000
Stddev	.0007	.0002	.0004	.0006	.0007
%RSD	186.9	55.06	119.1	67.03	10430.

#1	.0002	-.0002	-.0004	-.0005	-.0007
#2	-.0005	-.0007	.0003	-.0012	-.0007
#3	-.0009	-.0002	-.0003	-.0015	.0003
#4	-.0012	-.0004	-.0004	.0000	.0008
#5	.0005	-.0005	-.0009	-.0011	.0004

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.1147
Stddev	.1865
%RSD	2.0464

#1	9.3804
#2	9.2210
#3	9.0610
#4	8.9123
#5	8.9988

Method: ISOU Sample Name: icv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/08/99 13:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U [233]	U [234]	U [235]	U [236]	U [238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0005	.0008	.1462	.0022	20.53
Stddev	.0005	.0004	.0048	.0006	1.01
%RSD	86.54	42.98	3.309	24.97	4.908
#1	.0000	.0006	.1516	.0023	22.17
#2	-.0012	.0010	.1425	.0025	19.96
#3	-.0005	.0005	.1409	.0017	20.56
#4	-.0006	.0013	.1510	.0016	19.50
#5	-.0004	.0006	.1452	.0029	20.49

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.0266
Stddev	.2155
%RSD	2.3869

#1	8.8657
#2	9.1846
#3	9.0396
#4	8.7620
#5	9.2808

Method: ISOU Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/08/99 13:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0004	-.0008	-.0005	-.0003	-.0001
Stddev	.0004	.0003	.0003	.0004	.0005
%RSD	111.7	43.21	54.43	143.3	404.2

#1	-.0006	-.0011	-.0004	-.0009	.0006
#2	-.0004	-.0008	-.0010	-.0001	.0002
#3	-.0009	-.0005	-.0007	-.0004	-.0007
#4	-.0002	-.0010	-.0004	-.0001	-.0003
#5	.0002	-.0004	-.0002	.0001	-.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/s
Avg	9.0458
Stddev	.2803
%RSD	3.0990

#1	9.2231
#2	9.3611
#3	9.1396
#4	8.7912
#5	8.7143

HNF-1663 REV. 0

Method: ISOU

Sample Name: s98t002391

Operator: bjg

Comment: 101 df

Run Time: 03/08/99 13:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	.0003	.0003	.0007	.0003	.0006
Stddev	.0002	.0004	.0004	.0005	.0004
%RSD	65.55	127.0	63.11	162.1	73.82
#1	.0000	.0001	-.0006	-.0004	-.0005
#2	-.0004	-.0003	-.0004	-.0001	.0000
#3	-.0003	-.0005	-.0005	.0002	-.0005
#4	-.0006	.0000	-.0015	-.0002	-.0007
#5	-.0003	-.0008	-.0005	-.0012	-.0011

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.2041
Stddev	.2050
%RSD	2.2271

#1	8.8712
#2	9.3495
#3	9.2978
#4	9.3581
#5	9.1439

Method: ISOU

Sample Name: s98t002391_d

Operator: bjg

Comment: 101 df

Run Time: 03/08/99 13:50 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0007	-.0002	-.0005	-.0006	-.0004
Stddev	.0006	.0003	.0006	.0003	.0005
%RSD	89.50	153.1	120.8	42.10	125.3

#1	-.0004	-.0004	-.0011	-.0007	-.0012
#2	-.0003	.0001	-.0009	-.0006	-.0005
#3	-.0002	.0000	.0004	-.0002	-.0003
#4	-.0016	-.0002	-.0003	-.0006	-.0001
#5	-.0008	-.0005	-.0005	-.0010	.0001

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.2281
Stddev	.1658
%RSD	1.7962

#1	9.2167
#2	9.2820
#3	9.2429
#4	9.4285
#5	8.9703

Method: ISOU

Sample Name: ccv

Operator: bjg

Comment: iso-u lcs

Run Time: 03/08/99 13:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0007	.0011	.1469	.0022	21.29
Stddev	.0004	.0003	.0065	.0007	1.00
%RSD	57.59	32.07	4.448	32.03	4.676
#1	-.0004	.0015	.1580	.0027	22.85
#2	-.0012	.0011	.1457	.0026	20.74
#3	-.0010	.0009	.1448	.0023	20.22
#4	-.0003	.0006	.1453	.0027	21.51
#5	-.0005	.0013	.1406	.0010	21.13

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.0546
Stddev	.4159
%RSD	4.5934

#1	8.3501
#2	9.1573
#3	9.3916
#4	9.0521
#5	9.3221

Method: ISOU

Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/08/99 13:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	U_[233]	U_[234]	U_[235]	U_[236]	U_[238]
Line	233/pulse	234/pulse	235/pulse	236/pulse	238/pulse
Units	ppb	ppb	ppb	ppb	ppb
Avg	-.0012	-.0004	-.0006	-.0007	-.0001
Stddev	.0004	.0006	.0004	.0006	.0006
%RSD	35.05	159.1	61.79	80.32	876.7
#1	-.0017	.0003	-.0013	.0000	.0002
#2	-.0014	-.0010	-.0007	-.0004	.0007
#3	-.0011	.0000	-.0003	-.0005	.0001
#4	-.0006	-.0002	-.0007	-.0014	-.0008
#5	-.0014	-.0009	-.0003	-.0013	-.0005

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	9.4683
Stddev	.1895
%RSD	2.0010

#1	9.2952
#2	9.6881
#3	9.6523
#4	9.4018
#5	9.3038

LABCORE Completed Worklist Report for Worklist# 28488

Analyst: bjg

Instrument: ICPMS1

Book#: 15B59

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) Field Blank (Actinide)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	EMSAC-QC TH232	QC	1.00e-2	1.05e-02	105.000 % Recovery	
1	ICV	0	EMSAC-QC NP237	QC	9.880e-3	1.23e-02	124.494 % Recovery	
1	ICV	0	EMSAC-QC PU239	QC	1.00e-2	1.09e-02	109.000 % Recovery	
1	ICV	0	EMSAC-QC 241	QC	1.03e-2	1.16e-02	112.621 % Recovery	
2	ICB	0	EMSAC-QC PA231	QC	1	<4.72e-6	ug/mL	
2	ICB	0	EMSAC-QC TH229	QC	1	<3.66e-6	ug/mL	
2	ICB	0	EMSAC-QC TH230	QC	1	<5.69e-6	ug/mL	
2	ICB	0	EMSAC-QC TH232	QC	1	<5.45e-6	ug/mL	
2	ICB	0	EMSAC-QC NP237	QC	1	<6.50e-6	ug/mL	
2	ICB	0	EMSAC-QC PU239	QC	1	<6.07e-6	ug/mL	
2	ICB	0	EMSAC-QC PU240	QC	1	<2.24e-6	ug/mL	
2	ICB	0	EMSAC-QC PU242	QC	1	<3.68e-6	ug/mL	
2	ICB	0	EMSAC-QC 241	QC	1	<5.36e-6	ug/mL	
2	ICB	0	EMSAC-QC 243	QC	1	<4.64e-6	ug/mL	
2	ICB	0	EMSAC-QC 244	QC	1	<3.36e-6	ug/mL	
3	SAMPLE	S98T002391	D EMSAC-D1 PA231-D1	LIQUID	N/A	<4.763e-04	4.76e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 TH229-D1	LIQUID	N/A	<3.698e-04	3.70e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 TH230-D1	LIQUID	N/A	<5.752e-04	5.75e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 TH232-D1	LIQUID	N/A	<5.504e-04	5.50e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 NP237-D1	LIQUID	N/A	<6.562e-04	6.56e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 PU239-D1	LIQUID	N/A	<6.128e-04	6.13e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 PU240-D1	LIQUID	N/A	<2.267e-04	2.27e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 PU242-D1	LIQUID	N/A	<3.716e-04	3.72e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 241-D1	LIQUID	N/A	<5.411e-04	5.41e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 243-D1	LIQUID	N/A	<4.687e-04	4.69e-004	ug/mL
3	SAMPLE	S98T002391	D EMSAC-D1 244-D1	LIQUID	N/A	<3.391e-04	3.39e-004	ug/mL
4	DUP	S98T002391	D EMSAC-D1 PA231-D1	LIQUID	<4.76e-4	<4.76e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 TH229-D1	LIQUID	<3.70e-4	<3.70e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 TH230-D1	LIQUID	<5.75e-4	<5.75e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 TH232-D1	LIQUID	<5.50e-4	<5.50e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 NP237-D1	LIQUID	<6.56e-4	<6.56e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 PU239-D1	LIQUID	<6.13e-4	<6.13e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 PU240-D1	LIQUID	<2.27e-4	<2.27e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 PU242-D1	LIQUID	<3.72e-4	<3.72e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 241-D1	LIQUID	<5.41e-4	<5.41e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 243-D1	LIQUID	<4.69e-4	<4.69e-4	RPD	
4	DUP	S98T002391	D EMSAC-D1 244-D1	LIQUID	<3.39e-4	<3.39e-4	RPD	
5	CCV	0	EMSAC-QC TH232	QC	1.00e-2	1.07e-02	107.000 % Recovery	
5	CCV	0	EMSAC-QC NP237	QC	9.88e-3	1.26e-02	127.530 % Recovery	
5	CCV	0	EMSAC-QC PU239	QC	1.00e-2	1.12e-02	112.000 % Recovery	
5	CCV	0	EMSAC-QC 241	QC	1.03e-2	1.20e-02	116.505 % Recovery	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 28488

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
6 CCB	0		EMSAC-QC PA231	QC	1	<4.72e-6		ug/mL
6 CCB	0		EMSAC-QC TH229	QC	1	<3.66e-6		ug/mL
6 CCB	0		EMSAC-QC TH230	QC	1	<5.69e-6		ug/mL
6 CCB	0		EMSAC-QC TH232	QC	1	<5.45e-6		ug/mL
6 CCB	0		EMSAC-QC NP237	QC	1	<6.50e-6		ug/mL
6 CCB	0		EMSAC-QC PU239	QC	1	<6.07e-6		ug/mL
6 CCB	0		EMSAC-QC PU240	QC	1	<2.24e-6		ug/mL
6 CCB	0		EMSAC-QC PU242	QC	1	<3.68e-6		ug/mL
6 CCB	0		EMSAC-QC 241	QC	1	<5.36e-6		ug/mL
6 CCB	0		EMSAC-QC 243	QC	1	<4.64e-6		ug/mL
6 CCB	0		EMSAC-QC 244	QC	1	<3.36e-6		ug/mL

Final page for worklist# 28488

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

03/09/99
_____
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28488

Analyst: BJG Instrument: ICPMS1 Book# 15258Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) Field Blank (Actinide)

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV				@MSAC-QC	QC		
2	ICB				@MSAC-QC	QC		
3	SAMPLE	S98T002391	0	D	@MSAC-D1	LIQUID	98000436	BY-105 (2)
		Analytes Requested: 241-D1 , 243-D1 , 244-D1 , NP237-D1, PA231-D1, PU239-D1, PU240-D1, PU242-D1, TH229-D1, TH230-D1, TH232-D1						
4	DUP	S98T002391	0	D	@MSAC-D1	LIQUID		
5	CCV				@MSAC-QC	QC		
6	CCB				@MSAC-QC	QC		

Final page for worklist # 28488

Signature

Date

3-8-99

Signature

Date

S98T002391
2391-1-10-1
"D.F.
101
"

Data Entry Comments:

Uploaded + Validated by:

Saul M. Peng

03/09/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Analysis Report

HNF-1663 REV. 0

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/09/99 10:02 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	176800.
Stddev	1171.
%RSD	.6619

#1	177200.
#2	176800.
#3	174900.
#4	175600.
#5	175300.
#6	178100.
#7	178100.
#8	177600.
#9	177400.
#10	177500.

bjg

3-9-99

BY-105 Field Blank

588T00 2381

11-60-1

File # AC0309A.TPT

D.L.# 28488

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3185 TO 3203

Method: ACTINIDE Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/09/99 10:04 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5618	.5784	.5984	.6729	.5925
Stddev	.0342	.0548	.0555	.0460	.0467
%RSD	6.081	9.474	9.277	6.834	7.886
#1	.5968	.6682	.5589	.7354	.6262
#2	.5824	.5659	.6568	.6527	.5659
#3	.5122	.5325	.6504	.6220	.6260
#4	.5752	.5873	.5954	.7048	.5225
#5	.5422	.5382	.5303	.6499	.6220

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5595	.6090	.5719	.5733	.5437
Stddev	.0527	.0516	.0279	.0478	.0455
%RSD	9.428	8.470	4.887	8.347	8.370

#1	.5841	.6262	.5421	.6136	.4791
#2	.4874	.5701	.5659	.5453	.5618
#3	.5203	.6382	.6138	.6260	.5773
#4	.6076	.6683	.5833	.5711	.5144
#5	.5981	.5422	.5542	.5103	.5861

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.6411	.5668
Stddev	.0681	.0616
%RSD	10.62	10.86

#1	.7480	.6052
#2	.5742	.6320
#3	.5976	.5610
#4	.6238	.4698
#5	.6618	.5662

Sample Name: blank Run Time: 03/09/99 10:04

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	81.582
Stddev	1.635
%RSD	2.0036

#1	79.318
#2	80.694
#3	81.996
#4	82.296
#5	83.605

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.010 ppm cal

Run Time: 03/09/99 10:08 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5136	.5960	713.5	831.5	1.679
Stddev	.0257	.0254	2.3	5.2	.070
%RSD	4.999	4.257	.3289	.6295	4.192
#1	.4990	.6081	709.7	831.7	1.606
#2	.5357	.5625	713.9	834.0	1.711
#3	.4796	.6269	713.3	824.6	1.613
#4	.5414	.5788	714.7	828.6	1.774
#5	.5121	.6035	715.9	838.4	1.690

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	2.137	.8103	.5189	.5360	.5591
Stddev	.077	.0724	.0813	.0373	.0162
%RSD	3.594	8.938	15.67	6.954	2.897
#1	2.054	.8381	.3976	.5574	.5808
#2	2.246	.7654	.6046	.5893	.5549
#3	2.096	.7251	.5325	.5061	.5438
#4	2.106	.9148	.4854	.5004	.5452
#5	2.184	.8083	.5742	.5267	.5706

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5236	.5135
Stddev	.0522	.0471
%RSD	9.968	9.168
#1	.6042	.5886
#2	.5243	.4975
#3	.5023	.5136
#4	.5265	.4593
#5	.4608	.5084

Sample Name: cal std Run Time: 03/09/99 10:08

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	88.258
Stddev	2.132
%RSD	2.4154

#1	85.512
#2	87.105
#3	88.265
#4	89.271
#5	91.136

Method: ACTINIDE Sample Name: cal std

Operator: bjg

Comment: 0.100 ppm

Run Time: 03/09/99 10:12 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5945	1.371	7210.	8083.	12.79
Stddev	.0703	.100	71.	141.	.33
%RSD	11.82	7.313	.9865	1.747	2.547
#1	.6492	1.469	7120.	7889.	12.95
#2	.5854	1.456	7245.	8205.	12.88
#3	.5735	1.399	7193.	8221.	13.14
#4	.4930	1.245	7183.	7994.	12.70
#5	.6715	1.288	7310.	8105.	12.28

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	16.41	3.737	.5119	.5273	.5582
Stddev	.25	.150	.0303	.0665	.0616
%RSD	1.551	4.028	5.928	12.60	11.04
#1	16.41	3.826	.5114	.4461	.5549
#2	16.71	3.842	.5484	.5039	.6076
#3	16.61	3.611	.5032	.5550	.5550
#4	16.12	3.538	.4674	.5076	.6135
#5	16.19	3.865	.5292	.6241	.4598

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5202	.5139
Stddev	.0567	.0318
%RSD	10.89	6.191
#1	.5803	.5658
#2	.4594	.5224
#3	.4810	.4884
#4	.5003	.5003
#5	.5803	.4927

Sample Name: cal std Run Time: 03/09/99 10:12

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	90.918
Stddev	.849
%RSD	.93422

#1	91.912
#2	89.965
#3	90.092
#4	91.282
#5	91.337

Method: ACTINIDE Sample Name: icv

Operator: bjg

Comment: msact.lcs

Run Time: 03/09/99 10:24 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4818	.5750	748.0	5.235	1.647
Stddev	.0524	.0431	8.5	.157	.094
%RSD	10.87	7.504	1.130	2.996	5.702
#1	.5007	.5889	736.1	5.314	1.735
#2	.5353	.5424	746.9	5.462	1.750
#3	.5056	.6116	755.5	5.211	1.633
#4	.4697	.6145	756.7	5.089	1.582
#5	.3979	.5176	744.6	5.099	1.535

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	998.9	927.3	25.80	.7006	.5249
Stddev	6.0	11.4	.42	.0609	.0535
%RSD	.6003	1.235	1.630	8.692	10.18
#1	992.4	907.0	25.45	.6806	.5149
#2	999.0	931.7	26.22	.7466	.4649
#3	1007.	933.4	25.28	.7813	.6116
#4	993.7	930.1	26.17	.6533	.5227
#5	1003.	934.0	25.89	.6409	.5106

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1035.	.4945
Stddev	9.	.0508
%RSD	.8820	10.27
#1	1020.	.5783
#2	1037.	.4508
#3	1040.	.4631
#4	1036.	.4768
#5	1044.	.5036

Sample Name: icv Run Time: 03/09/99 10:24

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	94.501
Stddev	.163
%RSD	.17197

#1	94.525
#2	94.646
#3	94.287
#4	94.386
#5	94.659

Method: ACTINIDE Sample Name: icb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 10:38 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U_[238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4353	.5149	.6624	.6114	.5189
Stddev	.0434	.0104	.0266	.0562	.0576
%RSD	9.959	2.010	4.019	9.188	11.11

#1	.4195	.5006	.6980	.6592	.4583
#2	.4326	.5240	.6682	.5873	.4678
#3	.4996	.5135	.6488	.5239	.5898
#4	.4446	.5258	.6704	.6493	.5152
#5	.3802	.5104	.6266	.6371	.5632

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5015	.6042	.5235	.4988	.5115
Stddev	.0361	.0365	.0542	.0295	.0548
%RSD	7.202	6.034	10.35	5.906	10.72

#1	.4407	.6381	.5323	.4935	.5041
#2	.5240	.5768	.5803	.5451	.6049
#3	.4996	.5586	.4337	.4718	.4614
#4	.5117	.6069	.5293	.4764	.4870
#5	.5315	.6407	.5421	.5069	.4999

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.4602	.4671
Stddev	.0440	.0356
%RSD	9.569	7.618

#1	.4830	.5182
#2	.5100	.4678
#3	.4059	.4545
#4	.4799	.4199
#5	.4224	.4752

Sample Name: icb Run Time: 03/09/99 10:38

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	94.914
Stddev	.661
%RSD	.69681

#1	94.555
#2	94.776
#3	96.077
#4	94.465
#5	94.695

Method: ACTINIDE Sample Name: s98t002391

Operator: bjg

Comment: 101 df

Run Time: 03/09/99 10:44 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4808	.4700	.5918	.5808	.4426
Stddev	.0447	.0310	.0676	.0259	.0330
%RSD	9.292	6.586	11.42	4.458	7.463

#1	.4874	.5084	.5470	.5540	.4453
#2	.5386	.4287	.6484	.6236	.4854
#3	.4673	.4501	.6769	.5738	.4020
#4	.4158	.4817	.5198	.5717	.4609
#5	.4949	.4811	.5671	.5808	.4193

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4725	.5767	.5117	.5108	.5204
Stddev	.0504	.0440	.0402	.0341	.0497
%RSD	10.66	7.633	7.860	6.679	9.554

#1	.4137	.6311	.5505	.5365	.4839
#2	.4464	.5775	.4642	.5563	.4535
#3	.5360	.5773	.4914	.4776	.5395
#4	.5128	.5891	.4955	.4990	.5717
#5	.4536	.5086	.5567	.4846	.5533

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5079	.4859
Stddev	.0503	.0385
%RSD	9.904	7.934

#1	.4839	.5260
#2	.5563	.4323
#3	.4433	.4982
#4	.5613	.4609
#5	.4949	.5121

Sample Name: s98t002391 Run Time: 03/09/99 10:44

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	95.867
Stddev	1.277
%RSD	1.3321

#1	95.066
#2	94.076
#3	97.007
#4	96.197
#5	96.992

Method: ACTINIDE Sample Name: s98t002391_d

Operator: bjg

Comment: 101 df

Run Time: 03/09/99 10:48 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4892	.4766	.6161	.5450	.5232
Stddev	.0447	.0621	.0425	.0346	.0343
%RSD	9.143	13.04	6.892	6.347	6.551
#1	.4583	.5138	.5451	.5624	.5069
#2	.4965	.4261	.6409	.5951	.5353
#3	.5118	.4499	.6354	.5289	.5049
#4	.4327	.4258	.6093	.5331	.4916
#5	.5466	.5672	.6497	.5054	.5775

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5045	.5960	.4899	.5157	.5176
Stddev	.0401	.0703	.0267	.0353	.0195
%RSD	7.955	11.79	5.440	6.842	3.774
#1	.4583	.5867	.5034	.5381	.5103
#2	.5000	.6092	.4613	.5212	.5000
#3	.4740	.7041	.4671	.4568	.5015
#4	.5504	.5678	.5262	.5470	.5435
#5	.5397	.5122	.4916	.5157	.5329

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5117	.4853
Stddev	.0525	.0628
%RSD	10.27	12.94

#1	.5555	.4756
#2	.5458	.3909
#3	.5152	.5186
#4	.5193	.4812
#5	.4228	.5604

Sample Name: s98t002391_d Run Time: 03/09/99 10:48

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	96.194
Stddev	.965
%RSD	1.0031

#1	96.014
#2	94.658
#3	97.052
#4	96.285
#5	96.962

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccv

Operator: bjg

Comment: msact lcs

Run Time: 03/09/99 10:55 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5126	.5562	764.2	5.360	1.627
Stddev	.0526	.0197	10.2	.105	.097
%RSD	10.27	3.536	1.331	1.966	5.968
#1	.4900	.5530	766.8	5.201	1.600
#2	.5938	.5443	759.1	5.301	1.675
#3	.5140	.5668	759.8	5.432	1.644
#4	.4496	.5332	754.6	5.433	1.478
#5	.5156	.5837	780.6	5.432	1.737

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1023.	949.8	26.65	.7190	.5490
Stddev	9.	17.9	.91	.0535	.0337
%RSD	.8751	1.879	3.432	7.436	6.143
#1	1010.	930.4	26.20	.6440	.5425
#2	1030.	939.4	25.85	.7952	.5089
#3	1026.	954.6	26.26	.7181	.5280
#4	1016.	947.4	26.78	.7179	.5925
#5	1031.	977.4	28.18	.7198	.5729

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	1069.	.5173
Stddev	10.	.0550
%RSD	.9156	10.63
#1	1057.	.5250
#2	1063.	.4276
#3	1066.	.5104
#4	1073.	.5646
#5	1083.	.5586

Sample Name: ccv Run Time: 03/09/99 10:55

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	94.595
Stddev	.984
%RSD	1.0404

#1	95.236
#2	94.314
#3	94.691
#4	95.645
#5	93.087

HNF-1663 REV. 0

Method: ACTINIDE Sample Name: ccb

Operator: bjg

Comment: 2% hno3

Run Time: 03/09/99 11:02 Type: Unk

Mode: IR

Corr.Fact: 1.000000

Elem	Th[229]	Th[230]	Th[232]	U [238]	Pa[231]
Line	229/pulse	230/pulse	232/pulse	238/pulse	231/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.4588	.4613	.7018	.6108	.5040
Stddev	.0342	.0213	.0494	.0260	.0391
%RSD	7.465	4.617	7.037	4.258	7.751
#1	.4938	.4329	.7657	.6118	.4508
#2	.4323	.4571	.7015	.5811	.4960
#3	.4823	.4546	.7322	.6038	.4997
#4	.4720	.4720	.6421	.6525	.5136
#5	.4137	.4902	.6675	.6049	.5597

Elem	Np[237]	Pu[239]	Pu[240]	Pu[242]	Pu[244]
Line	237/pulse	239/pulse	240/pulse	242/pulse	244/pulse
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.5098	.7594	.4812	.4922	.4804
Stddev	.0399	.0375	.0264	.0546	.0353
%RSD	7.834	4.937	5.481	11.09	7.356
#1	.4902	.7299	.4795	.4473	.4544
#2	.5102	.7370	.5031	.5102	.5102
#3	.4546	.7704	.4997	.4303	.4685
#4	.5379	.7392	.4373	.5067	.4442
#5	.5562	.8204	.4867	.5666	.5249

Elem	Am[241]	Am[243]
Line	241/pulse	243/pulse
Units	Cts/S	Cts/S
Avg	.5329	.5270
Stddev	.0476	.0303
%RSD	8.927	5.751
#1	.4973	.5546
#2	.5421	.5527
#3	.4997	.5240
#4	.5136	.5241
#5	.6118	.4797

Sample Name: ccb Run Time: 03/09/99 11:02

Int. Std.	Ir[193]
Line	193/pulse
Units	Cts/S
Avg	95.047
Stddev	1.343
%RSD	1.4133

#1	93.161
#2	94.081
#3	96.060
#4	96.046
#5	95.887

LABCORE Completed Worklist Report for Worklist# 28544

Analyst: bjg

Instrument: ICPMS1

Book#: 990310-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (I129) DIRECT

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	I129	QC	0.00927	9.71e-03	104.746 % Recovery	
2 ICB	0	I129	QC	1	<1.20e-4	ug/mL	
3 SAMPLE	S98T002391 0 D	I129-D1	LIQUID	N/A <	1.212e-02	1.21e-002	ug/mL
4 DUP	S98T002391 0 D	I129-D1	LIQUID	<1.21e-2	<1.21e-2	RPD	
5 SAMPLE	S98T002822 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
6 DUP	S98T002822 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
7 SPK	S98T002822 0 D	I129-D1	LIQUID	n/a	6.12e-01	% Recovery	
8 SAMPLE	S98T002823 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
9 DUP	S98T002823 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
10 SAMPLE	S98T002824 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
11 DUP	S98T002824 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
12 CCV	0	I129	QC	0.00927	9.86e-03	106.365 % Recovery	
13 CCB	0	I129	QC	1	<1.20e-4	ug/mL	
14 SAMPLE	S98T002826 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
15 DUP	S98T002826 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
16 SAMPLE	S98T002827 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
17 DUP	S98T002827 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
18 SAMPLE	S98T002828 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
19 DUP	S98T002828 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
20 SAMPLE	S98T002829 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
21 DUP	S98T002829 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
22 SAMPLE	S98T002817 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860	ug/mL
23 DUP	S98T002817 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
24 CCV	0	I129	QC	0.00927	9.87e-03	106.472 % Recovery	
25 CCB	0	I129	QC	1	<1.20e-4	ug/mL	

Final page for worklist# 28544

Analyst Signature

Date

Analyst Signature

Date

Saul M. Pung 03/16/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28544

Analyst: BSJ Instrument: ICPMS1 Book# 99030-1Method: LA-506-101 Rev/Mod A.O

Worklist Comment: ICP/MS BY-105 (I129) DIRECT

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1 ICV				I129	QC		
2 ICB				I129	QC		
3 SAMPLE	S98T002391	0	D	I129-D1	LIQUID	98000436	BY-105 (2)
	Analytes Requested: I129-D1						
4 DUP	S98T002391	0	D	I129-D1	LIQUID		
5 SAMPLE	S98T002822	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
6 DUP	S98T002822	0	D	I129-D1	LIQUID		
7 SPK	S98T002822	0	D	I129-D1	LIQUID		
8 SAMPLE	S98T002823	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
9 DUP	S98T002823	0	D	I129-D1	LIQUID		
10 SAMPLE	S98T002824	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
11 DUP	S98T002824	0	D	I129-D1	LIQUID		
12 CCV				I129	QC		
13 CCB				I129	QC		
14 SAMPLE	S98T002826	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
15 DUP	S98T002826	0	D	I129-D1	LIQUID		
16 SAMPLE	S98T002827	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						

Data Entry Comments:

S98T002822

Post spike for I-129 was not done, instead I-127 was spiked.

$$I-127 = \frac{9.957 - 0}{10.06} \times 100 = 98.98\%$$

10.06

03/16/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3205

LABCORE Data Entry Template for Worklist# 28544

S Type	Sample#	R	A	Test	Matrix	Group#	Project
17 DUP	S98T002827	0	D	I129-D1	LIQUID		
18 SAMPLE	S98T002828	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
19 DUP	S98T002828	0	D	I129-D1	LIQUID		
20 SAMPLE	S98T002829	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
21 DUP	S98T002829	0	D	I129-D1	LIQUID		
22 SAMPLE	S98T002817	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
23 DUP	S98T002817	0	D	I129-D1	LIQUID		
24 CCV				I129	QC		
25 CCB				I129	QC		

Final page for worklist # 28544

2-12-98
Signature Date

_____
Signature Date

Data Entry Comments:

Uploaded & Validated by:
Saul M. Pang
03/16/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3206

03/09/99 15:50

WS2

Page: 1

LABCORE Data Entry Template for Worklist# 28544

Analyst: BTG Instrument: ICPMS1 Book# 990310-1Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (I129) DIRECT

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV			I129	QC		
2	ICB			I129	QC		
3	SAMPLE S98T002391	0	D	I129-D1	LIQUID	98000436	BY-105 (2)
	Analytes Requested: I129-D1						
4	DUP S98T002391	0	D	I129-D1	LIQUID		
~~5~~	~~SAMPLE S98T002391~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~		
6	SAMPLE S98T002822	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
7	DUP S98T002822	0	D	I129-D1	LIQUID		
→	~~SAMPLE S98T002822~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~		
8	SAMPLE S98T002823	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
9	DUP S98T002823	0	D	I129-D1	LIQUID		
10	SAMPLE S98T002824	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
11	DUP S98T002824	0	D	I129-D1	LIQUID		
12	CCV			I129	QC		
13	CCB			I129	QC		
~~14~~	~~SAMPLE S98T002825~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
~~15~~	~~DUP S98T002825~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~		
✓16	SAMPLE S98T002826	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3207

03/09/99 15:50

Page: 2

WS2

LABCORE Data Entry Template for Worklist# 28544

S Type	Sample#	R	A	Test	Matrix	Group#	Project
✓ 17 DUP	S98T002826	0	D	I129-D1	LIQUID		
18 SAMPLE	S98T002827	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
19 DUP	S98T002827	0	D	I129-D1	LIQUID		
✓ 20 SAMPLE	S98T002828	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
✓ 21 DUP	S98T002828	0	D	I129-D1	LIQUID		
22 SAMPLE	S98T002829	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
✓ 23 DUP	S98T002829	0	D	I129-D1	LIQUID		
→ 24 CCV	S98T002817			I129	QC		
	S98T002817			I129	QC		
25 CCB				I129	QC		

Final page for worklist # 28544

Signature *[Signature]* Date 3-12-99

S98T002891 .1-10m1
 2851-D .1-10m1
 2822 .1m1-10m1-.025m1-10m1
 2822-D
 2822-A
 2823
 2823-D
 2824
 2824-D
 2826
 2826-D
 2827
 2827-D
 2828
 2828-D
 2829
 2829-D
 2817
 2817-D

Data Entry Comments

Signature *[Signature]* Date

DF
 101
 101
 40501

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

POST SPIKE A = .020-1 of 54N21-E

3208

Method: MS_TUNE Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/12/99 08:56 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	183900.
Stddev	1195.
%RSD	.6498

#1	184600.
#2	182400.
#3	181600.
#4	183900.
#5	184900.
#6	183600.
#7	183300.
#8	183900.
#9	185000.
#10	185300.

Spence
3-12-99

64-101

F-129

598700 2391

2822

2823

2824

2826

2827

2828

2829

2817

FILE # 10312A.TXT

W.C.# 28544

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3209 TO 3238

Method: I-129

Sample Name: Blank

Operator:

Comment:

Run Time: 03/12/99 09:08 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	59.79	92.30	76.15
Stddev	2.25	.72	2.28
%RSD	3.756	.7824	2.994
#1	58.54	93.20	74.43
#2	61.08	91.57	78.35
#3	60.16	92.85	78.16
#4	56.69	91.65	73.20
#5	62.47	92.21	76.62

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.462
Stddev	.141
%RSD	.97572

#1	14.382
#2	14.562
#3	14.361
#4	14.659
#5	14.343

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	189.1016	59.7890	1.0000000	03/12/99 09:09:28
I_[129]	187.1562	92.2967	1.0000000	03/12/99 09:09:28
Xe[131]	-4.5573	76.1529	-1.0000000	03/12/99 09:09:28

HNF-1663 REV. 0

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/12/99 09:10 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3471.	3471.	3471.
Stddev	48.	48.	48.
%RSD	1.381	1.381	1.381

#1	3490.	3490.	3490.
#2	3400.	3400.	3400.
#3	3482.	3482.	3482.
#4	3453.	3453.	3453.
#5	3529.	3529.	3529.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.617
Stddev	.053
%RSD	.36108

#1	14.615
#2	14.647
#3	14.650
#4	14.646
#5	14.526

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	204.1238	59.7890	1.0000000	03/12/99 09:11:26
I_[129]	202.1784	92.2967	1.0000000	03/12/99 09:11:26
Xe[131]	203.1445	76.1529	1.0000000	03/12/99 09:11:26

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/12/99 09:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.003	9.708	.0427
Stddev	.048	.348	.0077
%RSD	2.408	3.589	17.97
#1	1.995	9.614	.0475
#2	1.954	9.266	.0356
#3	2.046	10.22	.0334
#4	2.059	9.629	.0466
#5	1.960	9.812	.0505

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.450
Stddev	.231
%RSD	2.2116
#1	10.781
#2	10.536
#3	10.152
#4	10.405
#5	10.375

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmaH

Run Time: 03/12/99 09:14 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0016	.0069	-.0043
Stddev	.0050	.0129	.0079
%RSD	304.0	187.5	182.8
#1	.0063	-.0147	-.0099
#2	.0009	.0050	-.0014
#3	-.0026	.0118	.0064
#4	-.0036	.0153	-.0140
#5	.0071	.0169	-.0028

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.560
Stddev	.110
%RSD	.75600

#1	14.554
#2	14.392
#3	14.537
#4	14.665
#5	14.653

Method: I-129

Sample Name: s98t002391

Operator: bjg

Comment: 101 df

Run Time: 03/12/99 09:17 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0190	-.0004	-.0067
Stddev	.0115	.0116	.0113
%RSD	60.36	2919.	167.9

#1	-.0083	.0146	-.0070
#2	-.0209	-.0120	-.0238
#3	-.0232	.0055	.0076
#4	-.0076	.0020	-.0073
#5	-.0351	-.0122	-.0032

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.047
Stddev	.110
%RSD	.73183

#1	15.086
#2	14.920
#3	14.987
#4	15.032
#5	15.212

Method: I-129

Sample Name: s98t002391_d

Operator: bjg

Comment: 101 df

Run Time: 03/12/99 09:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0177	-.0089	-.0005
Stddev	.0136	.0075	.0076
%RSD	77.01	83.55	1592.

#1	-.0315	-.0045	.0081
#2	-.0122	-.0120	-.0116
#3	-.0105	.0017	.0007
#4	-.0323	-.0171	.0042
#5	-.0017	-.0127	-.0038

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.971
Stddev	.074
%RSD	.49240

#1	15.031
#2	15.062
#3	14.956
#4	14.915
#5	14.890

HNF-1663 REV.0

Method: I-129

Sample Name: s98t002822

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0233	.0208	.0074
Stddev	.0098	.0103	.0111
%RSD	41.93	49.46	149.7

#1	.0134	.0210	-.0010
#2	.0306	.0172	.0069
#3	.0162	.0152	.0058
#4	.0200	.0123	-.0008
#5	.0363	.0382	.0261

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	14.965
Stddev	.169
%RSD	1.1271

#1	15.155
#2	15.036
#3	14.954
#4	14.983
#5	14.696

Method: I-129

Sample Name: 898t002822_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:24

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0224	.0192	-.0041
Stddev	.0098	.0099	.0070
%RSD	44.00	51.82	172.6
#1	.0213	.0126	-.0058
#2	.0243	.0343	.0055
#3	.0368	.0127	-.0111
#4	.0093	.0243	-.0095
#5	.0202	.0119	.0006

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.110
Stddev	.111
%RSD	.73476

#1	15.052
#2	15.011
#3	15.164
#4	15.044
#5	15.280

Method: I-129

Sample Name: s98t002822_a

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:26 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	9.957	.0151	.0054
Stddev	.116	.0097	.0095
%RSD	1.161	64.57	174.3
#1	9.867	.0022	-.0055
#2	9.883	.0115	.0006
#3	10.08	.0288	.0151
#4	10.09	.0185	.0012
#5	9.868	.0144	.0157

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.184
Stddev	.085
%RSD	.56153

#1	15.259
#2	15.234
#3	15.080
#4	15.104
#5	15.245

Method: I-129

Sample Name: s98t002823

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0129	.0354	.0067
Stddev	.0030	.0065	.0074
%RSD	23.29	18.47	110.9

#1	-.0147	.0313	.0095
#2	-.0101	.0314	.0100
#3	-.0170	.0354	.0075
#4	-.0127	.0467	.0126
#5	-.0100	.0323	-.0062

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.228
Stddev	.179
%RSD	1.1746

#1	14.957
#2	15.220
#3	15.429
#4	15.195
#5	15.341

Method: I-129   Sample Name: s98t002823_d   Operator: bjg
Comment: 40501 df
Run Time: 03/12/99 09:35 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0065	.0185	-.0104
Stddev	.0052	.0163	.0064
%RSD	80.01	87.99	61.63
#1	-.0068	.0265	-.0130
#2	-.0075	.0068	-.0120
#3	.0024	.0428	.0008
#4	-.0100	.0031	-.0122
#5	-.0106	.0133	-.0154

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.607
Stddev	.083
%RSD	.53477

#1	15.576
#2	15.715
#3	15.488
#4	15.638
#5	15.620

Method: I-129 Sample Name: s98t002824

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0147	.0123	-.0050
Stddev	.0078	.0115	.0132
%RSD	53.34	93.40	263.3

#1	-.0279	.0124	.0006
#2	-.0121	.0021	-.0242
#3	-.0106	.0267	.0086
#4	-.0078	.0204	-.0126
#5	-.0151	-.0001	.0025

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.635
Stddev	.154
%RSD	.98499

#1	15.668
#2	15.735
#3	15.564
#4	15.407
#5	15.799

Method: I-129

Sample Name: s98t002824_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:39 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0059	.0214	-.0011
Stddev	.0100	.0070	.0072
%RSD	168.9	32.42	647.5

#1	-.0167	.0201	-.0013
#2	-.0161	.0168	.0094
#3	.0022	.0310	-.0095
#4	.0044	.0258	-.0058
#5	-.0034	.0137	.0016

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.527
Stddev	.085
%RSD	.54872

#1	15.589
#2	15.514
#3	15.636
#4	15.425
#5	15.474

HNF-1663 REV.0

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/12/99 09:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.964	9.857	.0266
Stddev	.058	.272	.0074
%RSD	2.976	2.758	27.97
#1	1.929	9.705	.0252
#2	1.899	9.581	.0266
#3	2.025	10.24	.0383
#4	1.941	9.720	.0175
#5	2.027	10.04	.0255

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.296
Stddev	.271
%RSD	2.4030

#1	11.677
#2	11.427
#3	10.985
#4	11.284
#5	11.108

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmaH

Run Time: 03/12/99 09:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0051	-.0154	-.0103
Stddev	.0031	.0164	.0103
%RSD	60.40	106.4	99.99
#1	-.0052	.0015	-.0173
#2	-.0078	-.0298	.0065
#3	-.0006	-.0022	-.0193
#4	-.0082	-.0113	-.0086
#5	-.0038	-.0354	-.0126

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.165
Stddev	.077
%RSD	.50894

#1	15.104
#2	15.224
#3	15.069
#4	15.175
#5	15.251

Method: I-129

Sample Name: s98t002826

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0055	.0215	-.0088
Stddev	.0072	.0145	.0100
%RSD	131.3	67.34	113.7

#1	-.0079	.0065	-.0224
#2	-.0150	.0264	-.0051
#3	-.0009	.0069	.0046
#4	.0038	.0282	-.0077
#5	-.0075	.0397	-.0135

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.506
Stddev	.105
%RSD	.67590

#1	15.571
#2	15.435
#3	15.521
#4	15.634
#5	15.372

Method: I-129

Sample Name: s98t002826_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0042	.0178	-.0143
Stddev	.0094	.0138	.0077
%RSD	222.8	77.71	54.09
#1	.0034	.0255	-.0086
#2	-.0105	.0329	-.0104
#3	-.0173	.0012	-.0236
#4	-.0015	.0052	-.0072
#5	..0047	.0243	-.0216

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.703
Stddev	.091
%RSD	.57786

#1	15.771
#2	15.619
#3	15.826
#4	15.637
#5	15.663

Method: I-129

Sample Name: s98t002828 *2/28 3-12-99*

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:51 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ ₁₂₇	I_ ₁₂₉	Xe ₁₃₁
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0020	.0182	-.0141
Stddev	.0077	.0087	.0074
%RSD	381.5	47.60	52.52

#1	.0090	.0125	-.0119
#2	-.0077	.0151	-.0201
#3	-.0045	.0103	-.0177
#4	.0089	.0319	-.0021
#5	.0044	.0214	-.0184

Int. Std.	In ₁₁₅
Line	115/pulse
Units	Cts/S
Avg	15.727
Stddev	.132
%RSD	.83640

#1	15.740
#2	15.757
#3	15.925
#4	15.607
#5	15.607

Method: I-129   Sample Name: s98t002827_d   Operator: bjg
Comment: 40501 df
Run Time: 03/12/99 09:54   Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0010	.0289	-.0003
Stddev	.0079	.0089	.0175
%RSD	797.6	30.88	6621.
#1	.0067	.0178	-.0029
#2	-.0115	.0395	-.0177
#3	.0087	.0222	-.0137
#4	.0001	.0349	.0258
#5	.0011	.0300	.0072

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.680
Stddev	.194
%RSD	1.2355
#1	15.918
#2	15.801
#3	15.416
#4	15.584
#5	15.682

Method: I-129

Sample Name: s98t002828

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

09:56 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0137	.0225	.0001
Stddev	.0093	.0153	.0089
%RSD	68.23	67.96	6333.

#1	-.0223	.0088	-.0093
#2	-.0024	.0039	.0067
#3	-.0245	.0322	.0060
#4	-.0094	.0388	.0072
#5	-.0099	.0289	-.0099

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.906
Stddev	.070
%RSD	.44205

#1	15.966
#2	15.895
#3	15.986
#4	15.871
#5	15.814

Method: I-129

Sample Name: s98t002828_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 09:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0045	.0216	-.0056
Stddev	.0097	.0141	.0086
%RSD	217.4	65.05	153.2

#1	-.0124	.0102	-.0010
#2	-.0028	.0292	-.0209
#3	-.0160	.0122	-.0025
#4	.0010	.0136	-.0011
#5	.0078	.0428	-.0026

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	15.757
Stddev	.242
%RSD	1.5366

#1	16.076
#2	15.839
#3	15.844
#4	15.526
#5	15.502

Method: I-129 Sample Name: s98t002829

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 10:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0096	.0283	-.0020
Stddev	.0070	.0143	.0112
%RSD	73.19	50.55	565.9
#1	-.0172	.0152	-.0148
#2	-.0040	.0381	-.0018
#3	-.0006	.0314	.0091
#4	-.0142	.0449	.0091
#5	-.0121	.0120	-.0114

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.751
Stddev	.207
%RSD	1.3149
#1	15.949
#2	15.641
#3	15.444
#4	15.826
#5	15.894

Method: I-129

Sample Name: s98t002829_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

10:03 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0037	.0292	-.0036
Stddev	.0079	.0141	.0045
%RSD	212.9	48.13	126.8

#1	-.0038	.0244	.0015
#2	-.0049	.0324	-.0041
#3	.0033	.0374	-.0078
#4	-.0163	.0076	-.0082
#5	.0030	.0442	.0007

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.813
Stddev	.183
%RSD	1.1544

#1	15.689
#2	15.879
#3	15.762
#4	16.097
#5	15.639

Method: I-129 Sample Name: s98t002817

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 10:05 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0292	.0389	-.0026
Stddev	.0055	.0074	.0096
%RSD	19.00	19.01	375.4

#1	.0315	.0327	-.0061
#2	.0275	.0419	-.0105
#3	.0297	.0308	-.0098
#4	.0211	.0491	.0125
#5	.0362	.0397	.0011

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.772
Stddev	.140
%RSD	.88557

#1	15.700
#2	15.779
#3	15.601
#4	15.978
#5	15.803

Method: I-129

Sample Name: s98t002817_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 10:07

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0279	.0364	-.0046
Stddev	.0078	.0069	.0093
%RSD	28.07	19.00	202.8
#1	.0177	.0321	-.0139
#2	.0224	.0447	.0099
#3	.0300	.0412	-.0111
#4	.0373	.0366	-.0059
#5	.0321	.0274	-.0021

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.763
Stddev	.162
%RSD	1.0258

#1	15.838
#2	15.971
#3	15.792
#4	15.549
#5	15.668

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/12/99 10:09 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.990	9.865	.0353
Stddev	.032	.238	.0173
%RSD	1.618	2.411	48.99

#1	1.939	9.492	.0106
#2	2.023	10.04	.0479
#3	1.984	9.777	.0553
#4	1.997	9.948	.0314
#5	2.009	10.07	.0314

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.413
Stddev	.329
%RSD	2.8858

#1	11.977
#2	11.218
#3	11.145
#4	11.378
#5	11.347

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/12/99 10:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0053	-.0144	-.0123
Stddev	.0083	.0169	.0067
%RSD	156.3	117.3	54.57

#1	-.0082	-.0018	-.0082
#2	-.0174	-.0089	-.0152
#3	.0016	-.0438	-.0026
#4	-.0056	-.0116	-.0169
#5	.0032	-.0058	-.0187

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.714
Stddev	.069
%RSD	.43694

#1	15.728
#2	15.645
#3	15.755
#4	15.799
#5	15.643

LABCORE Completed Worklist Report for Worklist# 28545

Analyst: bjg

Instrument: ICPMS1

Book#: 990312-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (I129) DIRECT

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	I129	QC	0.00927	9.73e-03	104.962 % Recovery	
2 ICV	0	I129	QC	1	<1.20e-4	ug/mL	
3 SAMPLE	S98T002940 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
4 DUP	S98T002940 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
5 SAMPLE	S98T002965 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
6 DUP	S98T002965 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
7 SAMPLE	S98T002966 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
8 DUP	S98T002966 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
9 SAMPLE	S98T002967 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
10 DUP	S98T002967 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
11 SAMPLE	S98T002969 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
12 DUP	S98T002969 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
13 CCV	0	I129	QC	0.00927	9.83e-03	106.041 % Recovery	
14 CCB	0	I129	QC	1	<1.20e-4	ug/mL	
15 SAMPLE	S98T003010 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
16 DUP	S98T003010 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
17 SAMPLE	S98T003011 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
18 DUP	S98T003011 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
19 SAMPLE	S98T003012 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
20 DUP	S98T003012 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
21 SAMPLE	S98T003013 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
22 DUP	S98T003013 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
23 SAMPLE	S98T003385 0 D	I129-D1	LIQUID	N/A <	4.860e+00	4.860 ug/mL	
24 DUP	S98T003385 0 D	I129-D1	LIQUID	<4.86e0	<4.86e0	RPD	
25 CCV	0	I129	QC	0.00927	9.80e-03	105.717 % Recovery	
26 CCB	0	I129	QC	1	<1.20e-4	ug/mL	

Final page for worklist# 28545

Analyst Signature

Date

Analyst Signature

Date

Paul M. Pang 03/16/99
Reviewer Signature Date

03/09/99 16:08
ws2

FILE # 10312A.TPT

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 28545

Analyst: BJ6 Instrument: ICPMS1 Book# 890312-1

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (I129) DIRECT

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	ICV				I129	QC		
2	ICB				I129	QC		
3	SAMPLE	S98T002940	0	D	I129-D1	LIQUID	98000490	BY-105 (2)
		Analytes Requested: I129-D1						
4	DUP	S98T002940	0	D	I129-D1	LIQUID		
5	SAMPLE	S98T002965	0	D	I129-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: I129-D1						
✓ 6	DUP	S98T002965	0	D	I129-D1	LIQUID		
✓ 7	SAMPLE	S98T002966	0	D	I129-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: I129-D1						
✓ 8	DUP	S98T002966	0	D	I129-D1	LIQUID		
9	SAMPLE	S98T002967	0	D	I129-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: I129-D1						
✓ 10	DUP	S98T002967	0	D	I129-D1	LIQUID		
~~11~~	~~SAMPLE~~	~~S98T002968~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~	98000509	BY-105 (2)
		~~Analytes Requested: I129-D1~~						
~~12~~	~~DUP~~	~~S98T002968~~	~~0~~	~~D~~	~~I129-D1~~	~~LIQUID~~		
13	CCV				I129	QC		
14	CCB				I129	QC		
15	SAMPLE	S98T002969	0	D	I129-D1	LIQUID	98000509	BY-105 (2)
		Analytes Requested: I129-D1						
16	DUP	S98T002969	0	D	I129-D1	LIQUID		

Data Entry Comments:

Validated & Uploaded by:

S. M. Pang

03/16/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

ws2

LABCORE Data Entry Template for Worklist# 28545

S Type	Sample#	R	A	Test	Matrix	Group#	Project
17 SAMPLE	S98T003010	0	D	I129-D1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: I129-D1						
18 DUP	S98T003010	0	D	I129-D1	LIQUID		
19 SAMPLE	S98T003011	0	D	I129-D1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: I129-D1						
20 DUP	S98T003011	0	D	I129-D1	LIQUID		
21 SAMPLE	S98T003012	0	D	I129-D1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: I129-D1						
22 DUP	S98T003012	0	D	I129-D1	LIQUID		
23 SAMPLE	S98T003013	0	D	I129-D1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: I129-D1						
24 DUP	S98T003013	0	D	I129-D1	LIQUID		
✓25 SAMPLE	S98T003385	0	D	I129-D1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: I129-D1						
✓26 DUP	S98T003385	0	D	I129-D1	LIQUID		
27 CCV				I129	QC		
28 CCB				I129	QC		

Final page for worklist # 28545

3-12-98
Signature Date

Signature Date

S98T002840, 2865, 2866, 2867, 2868, 2869, 3010, 3011, 3012, 3013, 3385

Sampled + Dupl

1-10-1 - 1025-10-1

Dr. 40501

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/12/99 13:25 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	168800.
Stddev	1713.
%RSD	1.015
#1	168700.
#2	170200.
#3	166300.
#4	167300.
#5	170500.
#6	166200.
#7	169300.
#8	168400.
#9	170500.
#10	170500.

3-12-99

B4-105 I-128

598700 2940

2965

2966

2967

2968

2969

3010

3011

3012

3013

3385

File # 10312A.TRT

w.l.# 28545

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3242 TO 3248.

Method: I-129

Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/12/99 13:32 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ ¹²⁷	I_ ¹²⁹	Xe ¹³¹
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.942	9.731	.0771
Stddev	.034	.128	.0153
%RSD	1.742	1.317	19.90
#1	1.956	9.706	.0817
#2	1.949	9.824	.0795
#3	1.978	9.789	.0990
#4	1.887	9.517	.0597
#5	1.938	9.817	.0656

Int. Std.	In ¹¹⁵
Line	115/pulse
Units	Cts/s
Avg	9.8101
Stddev	.2896
%RSD	2.9523

#1	10.237
#2	9.9622
#3	9.6965
#4	9.6448
#5	9.5100

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmah

Run Time: 03/12/99 13:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0468	.0410	.0349
Stddev	.0064	.0085	.0054
%RSD	13.70	20.69	15.52

#1	-.0392	.0551	.0384
#2	-.0455	.0399	.0276
#3	-.0437	.0411	.0310
#4	-.0492	.0336	.0370
#5	-.0562	.0352	.0406

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.789
Stddev	.089
%RSD	.64761

#1	13.661
#2	13.791
#3	13.771
#4	13.811
#5	13.910

Method: I-129 Sample Name: s98t002940

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0017	.0691	.0262
Stddev	.0084	.0202	.0044
%RSD	486.5	29.25	16.87

#1	-.0079	.0451	.0218
#2	.0152	.0588	.0220
#3	.0014	.0945	.0257
#4	.0005	.0620	.0300
#5	-.0005	.0852	.0314

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.865
Stddev	.129
%RSD	.92969

#1	14.003
#2	14.008
#3	13.776
#4	13.751
#5	13.788

Method: I-129

Sample Name: s98t002940_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0070	.0701	.0207
Stddev	.0066	.0093	.0115
%RSD	93.73	13.27	55.74
#1	.0146	.0703	.0201
#2	-.0016	.0706	.0198
#3	.0024	.0784	.0109
#4	.0109	.0547	.0126
#5	.0087	.0763	.0399

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.856
Stddev	.140
%RSD	1.0075
#1	13.947
#2	14.019
#3	13.671
#4	13.764
#5	13.878

Method: I-129 Sample Name: s98t002965

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0116	.0540	.0267
Stddev	.0083	.0177	.0068
%RSD	71.87	32.74	25.41

#1	-.0228	.0584	.0217
#2	-.0089	.0491	.0308
#3	-.0173	.0474	.0345
#4	-.0022	.0336	.0179
#5	-.0066	.0814	.0287

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.868
Stddev	.145
%RSD	1.0433

#1	14.063
#2	13.773
#3	13.792
#4	13.981
#5	13.734

Method: I-129

Sample Name: s98t002965_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0179	.0620	.0226
Stddev	.0071	.0099	.0070
%RSD	39.52	15.90	30.72
#1	-.0183	.0618	.0162
#2	-.0230	.0770	.0276
#3	-.0177	.0540	.0322
#4	-.0242	.0649	.0171
#5	-.0063	.0525	.0202

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	13.814
Stddev	.168
%RSD	1.2193
#1	13.953
#2	13.712
#3	14.011
#4	13.602
#5	13.795

Method: I-129 Sample Name: s98t002966

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:48 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0095	.0647	.0328
Stddev	.0068	.0224	.0080
%RSD	71.15	34.66	24.48
#1	-.0013	.0373	.0217
#2	-.0041	.0977	.0433
#3	-.0101	.0683	.0367
#4	-.0166	.0678	.0301
#5	-.0156	.0526	.0321

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	13.963
Stddev	.274
%RSD	1.9602

#1	14.375
#2	13.637
#3	13.840
#4	13.920
#5	14.044

Method: I-129

Sample Name: s98t002966_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

13:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0071	.0571	.0265
Stddev	.0065	.0123	.0064
%RSD	91.30	21.61	24.09

#1	-.0009	.0693	.0272
#2	-.0172	.0472	.0290
#3	-.0095	.0696	.0260
#4	-.0057	.0567	.0338
#5	-.0024	.0428	.0164

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.013
Stddev	.169
%RSD	1.2093

#1	14.203
#2	13.914
#3	13.777
#4	14.123
#5	14.047

Method: I-129

Sample Name: s98t002967

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

13:52 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0072	.0561	.0281
Stddev	.0096	.0159	.0058
%RSD	132.7	28.42	20.63
#1	-.0057	.0531	.0272
#2	-.0238	.0449	.0270
#3	-.0052	.0378	.0194
#4	-.0004	.0759	.0329
#5	-.0009	.0687	.0339

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.003
Stddev	.194
%RSD	1.3844
#1	14.311
#2	13.774
#3	13.947
#4	13.987
#5	13.994

Method: I-129

Sample Name: s98t002967_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

13:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0071	.0622	.0235
Stddev	.0068	.0211	.0079
%RSD	96.36	33.91	33.51
#1	-.0050	.0490	.0166
#2	-.0095	.0515	.0171
#3	-.0137	.0993	.0238
#4	.0037	.0596	.0239
#5	-.0109	.0517	.0360

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.164
Stddev	.117
%RSD	.82886

#1	14.173
#2	14.099
#3	14.039
#4	14.351
#5	14.160

Method: I-129 Sample Name: s98t002969

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 13:58 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0023	.0459	.0065
Stddev	.0091	.0160	.0071
%RSD	395.8	34.86	107.8
#1	.0010	.0369	.0139
#2	-.0026	.0236	.0064
#3	.0043	.0497	-.0022
#4	-.0178	.0649	.0016
#5	.0036	.0544	.0132

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/s
Avg	14.491
Stddev	.154
%RSD	1.0596

#1	14.335
#2	14.623
#3	14.314
#4	14.567
#5	14.615

Method: I-129

Sample Name: s98t002969_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 14:00 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0063	.0298	.0013
Stddev	.0131	.0127	.0148
%RSD	208.6	42.74	1136.
#1	.0222	.0450	-.0069
#2	-.0078	.0273	.0135
#3	.0127	.0347	-.0011
#4	.0111	.0318	.0186
#5	-.0068	.0102	-.0176

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.564
Stddev	.180
%RSD	1.2367

#1	14.379
#2	14.587
#3	14.691
#4	14.383
#5	14.779

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/12/99 14:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.983	9.826	.0761
Stddev	.051	.117	.0109
%RSD	2.594	1.188	14.30

#1	2.032	9.821	.0728
#2	1.979	9.865	.0861
#3	1.913	9.672	.0590
#4	1.957	9.779	.0786
#5	2.033	9.990	.0841

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.542
Stddev	.158
%RSD	1.4966

#1	10.342
#2	10.529
#3	10.733
#4	10.659
#5	10.445

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/12/99 14:06 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0368	.0319	.0142
Stddev	.0066	.0083	.0097
%RSD	17.88	26.01	68.20

#1	-.0348	.0234	.0265
#2	-.0273	.0253	.0113
#3	-.0438	.0386	.0218
#4	-.0422	.0425	.0035
#5	-.0361	.0300	.0077

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.545
Stddev	.106
%RSD	.73024

#1	14.398
#2	14.505
#3	14.525
#4	14.656
#5	14.641

Method: I-129

Sample Name: S98t003010

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

14:08 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0130	.0492	.0174
Stddev	.0097	.0106	.0081
%RSD	74.64	21.62	46.18
#1	-.0287	.0515	.0140
#2	-.0086	.0499	.0230
#3	-.0027	.0654	.0222
#4	-.0110	.0384	.0233
#5	-.0142	.0409	.0048

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.932
Stddev	.305
%RSD	2.0447

#1	15.458
#2	14.811
#3	14.803
#4	14.910
#5	14.677

Method: I-129   Sample Name: s98t003010_d   Operator: bjg
Comment: 40501 df
Run Time: 03/12/99 14:10 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0022	.0583	.0211
Stddev	.0075	.0160	.0107
%RSD	344.9	27.37	50.56

#1	-.0048	.0704	.0221
#2	.0049	.0699	.0137
#3	-.0063	.0493	.0345
#4	-.0109	.0676	.0275
#5	.0063	.0344	.0078

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	14.886
Stddev	.072
%RSD	.48638

#1	14.889
#2	14.861
#3	14.819
#4	14.853
#5	15.008

Method: I-129 Sample Name: s98t003011

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 14:12 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0119	.0486	.0204
Stddev	.0025	.0068	.0119
%RSD	20.80	13.98	58.69

#1	-.0136	.0388	.0258
#2	-.0154	.0480	.0346
#3	-.0109	.0579	.0020
#4	-.0094	.0485	.0207
#5	-.0103	.0497	.0187

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.146
Stddev	.186
%RSD	1.2281

#1	15.469
#2	15.136
#3	15.053
#4	15.059
#5	15.014

Method: I-129

Sample Name: s98t003011_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

14:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0070	.0539	.0184
Stddev	.0084	.0099	.0072
%RSD	120.6	18.46	38.93
#1	-.0093	.0693	.0202
#2	-.0104	.0541	.0151
#3	.0078	.0471	.0103
#4	-.0095	.0555	.0295
#5	-.0136	.0434	.0170

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.108
Stddev	.148
%RSD	.98057

#1	15.028
#2	14.941
#3	15.062
#4	15.187
#5	15.320

Method: I-129 Sample Name: s98t003012

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 14:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0191	.0875	.0383
Stddev	.0212	.0439	.0191
%RSD	110.8	50.20	49.94
#1	-.0547	.1075	.0354
#2	-.0229	.1551	.0717
#3	-.0065	.0496	.0257
#4	-.0053	.0694	.0327
#5	-.0061	.0560	.0259

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.518
Stddev	.953
%RSD	6.1444

#1	17.211
#2	15.257
#3	15.132
#4	14.940
#5	15.051

Method: I-129   Sample Name: s98t003012_d   Operator: bjg
Comment: 40501 df
Run Time: 03/12/99 14:16 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0092	.0539	.0186
Stddev	.0076	.0037	.0063
%RSD	81.68	6.871	33.85

#1	-.0100	.0533	.0175
#2	-.0157	.0545	.0112
#3	-.0124	.0513	.0266
#4	-.0118	.0504	.0145
#5	.0037	.0598	.0232

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	15.146
Stddev	.078
%RSD	.51793

#1	15.231
#2	15.033
#3	15.204
#4	15.153
#5	15.111

Method: I-129

Sample Name: s98t003013

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

14:18 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0030	.0501	.0065
Stddev	.0043	.0120	.0177
%RSD	142.8	24.02	273.7
#1	-.0061	.0424	-.0042
#2	.0011	.0372	-.0117
#3	.0004	.0466	.0052
#4	-.0088	.0672	.0349
#5	-.0017	.0572	.0081

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.337
Stddev	.372
%RSD	2.4239

#1	15.970
#2	15.369
#3	15.159
#4	15.088
#5	15.097

Method: I-129

Sample Name: s98t003013_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

14:19 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0037	.0565	.0165
Stddev	.0077	.0126	.0092
%RSD	206.4	22.28	56.00
#1	-.0170	.0392	.0212
#2	-.0022	.0485	.0118
#3	-.0024	.0712	.0244
#4	.0026	.0613	.0025
#5	.0003	.0622	.0226

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.324
Stddev	.027
%RSD	.17848

#1	15.350
#2	15.327
#3	15.311
#4	15.284
#5	15.346

Method: I-129 Sample Name: s98t003385

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99 14:21 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0176	.0474	.0162
Stddev	.0037	.0070	.0071
%RSD	20.81	14.86	43.62

#1	-.0194	.0406	.0044
#2	-.0198	.0412	.0185
#3	-.0118	.0458	.0234
#4	-.0208	.0544	.0178
#5	-.0163	.0553	.0169

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.488
Stddev	.100
%RSD	.64811

#1	15.621
#2	15.479
#3	15.370
#4	15.418
#5	15.549

Method: I-129

Sample Name: s98t003385_d

Operator: bjg

Comment: 40501 df

Run Time: 03/12/99

14:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0180	.0461	.0178
Stddev	.0063	.0065	.0095
%RSD	35.08	14.19	53.64
#1	-.0146	.0464	.0114
#2	-.0281	.0566	.0262
#3	-.0144	.0414	.0296
#4	-.0201	.0399	.0080
#5	-.0128	.0463	.0137

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.431
Stddev	.115
%RSD	.74666

#1	15.294
#2	15.441
#3	15.451
#4	15.603
#5	15.365

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/12/99 14:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.948	9.802	.0484
Stddev	.056	.223	.0111
%RSD	2.897	2.273	22.88

#1	1.949	9.665	.0486
#2	1.857	9.516	.0333
#3	1.948	9.920	.0423
#4	2.006	9.820	.0602
#5	1.980	10.09	.0577

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.277
Stddev	.190
%RSD	1.6885

#1	11.491
#2	11.433
#3	11.264
#4	11.168
#5	11.026

Method: I-129

Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/12/99 14:26 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0408	.0285	.0239
Stddev	.0111	.0084	.0087
%RSD	27.17	29.61	36.23
#1	-.0396	.0240	.0250
#2	-.0349	.0366	.0362
#3	-.0600	.0386	.0128
#4	-.0367	.0222	.0262
#5	-.0325	.0210	.0194

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	15.311
Stddev	.210
%RSD	1.3717

#1	15.184
#2	15.011
#3	15.408
#4	15.420
#5	15.532

LABCORE Completed Worklist Report for Worklist# 28546

Analyst: bjg

Instrument: ICPMS1

Book#: 990919-3

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (I129) Solid Acid Digest

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 ICV	0	I129	QC	0.00927	9.88e-03	106.580	% Recovery
2 ICB	0	I129	QC	1	<1.20e-4		ug/mL
3 BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3		ug/g
4 SAMPLE	S98T002604 0 A	I129-A1	SOLID	N/A	4.444e+00	4.444	ug/g
5 DUP	S98T002604 0 A	I129-A1	SOLID	<4.44e0	<4.31e0		RPD
6 SPK	S98T002604 0 A	I129-A1	SOLID	N/A	2.40e-01		% Recovery
7 SAMPLE	S98T002605 0 A	I129-A1	SOLID	N/A	4.283e+00	4.283	ug/g
8 DUP	S98T002605 0 A	I129-A1	SOLID	<4.28e0	<4.42e0		RPD
9 SAMPLE	S98T002606 0 A	I129-A1	SOLID	N/A	4.522e+00	4.523	ug/g
10 DUP	S98T002606 0 A	I129-A1	SOLID	<4.52e0	<4.45e0		RPD
11 CCV	0	I129	QC	0.00927	9.83e-03	106.041	% Recovery
12 CCB	0	I129	QC	1	<1.20e-4		ug/mL

Final page for worklist# 28546

Analyst Signature Date

Analyst Signature Date

Sam M. Pang 03/23/99
Reviewer Signature Date

Post spike for I-129 was not done, instead I-127 was spiked.

S98T002604

$$I-127 = \frac{10.11 - 0}{10.06} \times 100 = 100.5\%$$

LABCORE Data Entry Template for Worklist# 28546

Analyst: BTE Instrument: ICPMS1 Book# 990319-3Method: LA-506-101 Rev/Mod AO

Worklist Comment: ICP/MS BY-105 (I129) Solid Acid Digest

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		I129	QC		
2	ICB		I129	QC		
3	BLNK-PREP		I129-A1	SOLID		
4	SAMPLE	S98T002604 0 A	I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
5	DUP	S98T002604 0 A	I129-A1	SOLID		
6	SPK	S98T002604 0 A	I129-A1	SOLID		
7	SAMPLE	S98T002605 0 A	I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
8	DUP	S98T002605 0 A	I129-A1	SOLID		
9	SAMPLE	S98T002606 0 A	I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
10	DUP	S98T002606 0 A	I129-A1	SOLID		
11	CCV		I129	QC		
12	CCB		I129	QC		

Final page for worklist # 28546

[Signature] 312-88
Signature Date
S98T002604 2604-10
2605
2605-10
2606
2606-10
Data Entry Comments:

1ml-9-1
105-20-1

26
10
201

Signature Date

Uploaded + Validated by:
[Signature]

HNF-1663 REV. 0

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/19/99

11:20 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	119900.
Stddev	485.
%RSD	.4046

#1	119300.
#2	120000.
#3	120100.
#4	120300.
#5	119500.
#6	119000.
#7	119900.
#8	120400.
#9	119700.
#10	120400.

3-19-99

B4-105 F-128

SS8T002809

2968

2604

2605

2606

File # 10315A-TV8

W.C. # 28848

28546

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
CALIBRATION/ANALYSIS ON PAGES 3271 TO 3287

Method: I-129 Sample Name: Blank

Operator:

Comment:

Run Time: 03/19/99 11:22 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	69.81	77.25	63.80
Stddev	2.04	1.68	2.63
%RSD	2.927	2.178	4.116
#1	71.01	78.72	66.98
#2	70.49	75.97	65.54
#3	67.24	78.41	64.09
#4	68.15	78.21	61.83
#5	72.14	74.96	60.57

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.7805
Stddev	.1157
%RSD	1.3171

#1	8.6890
#2	8.6815
#3	8.7299
#4	8.9440
#5	8.8581

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	204.2119	69.8058	1.0000000	03/19/99 11:23:22
I_[129]	203.7662	77.2529	1.0000000	03/19/99 11:23:22
Xe[131]	204.5712	63.8018	1.0000000	03/19/99 11:23:22

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/19/99 11:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3550.	3550.	3550.
Stddev	45.	45.	45.
%RSD	1.257	1.257	1.257

#1	3493.	3493.	3493.
#2	3565.	3565.	3565.
#3	3602.	3602.	3602.
#4	3514.	3514.	3514.
#5	3574.	3574.	3574.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.8265
Stddev	.0536
%RSD	.60748

#1	8.8228
#2	8.8951
#3	8.7451
#4	8.8369
#5	8.8328

HNF-1663 REV. 0

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.2443	69.8058	1.0000000	03/19/99 11:26:30
I_[129]	207.7987	77.2529	1.0000000	03/19/99 11:26:30
Xe[131]	208.6036	63.8018	1.0000000	03/19/99 11:26:30

3275

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/19/99 11:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.058	9.883	.0527
Stddev	.103	.501	.0301
%RSD	4.987	5.067	57.09

#1	1.990	9.476	.0301
#2	1.968	9.525	.0243
#3	2.037	9.947	.0555
#4	2.228	10.71	.1006
#5	2.069	9.753	.0529

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	6.7916
Stddev	.5068
%RSD	7.4625

#1	7.3704
#2	7.1077
#3	6.8710
#4	6.0684
#5	6.5403

Method: I-129

Sample Name: icb

Operator: bjg

Comment: 2% tmah

Run Time: 03/19/99 11:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0128	.0022	-.0014
Stddev	.0033	.0076	.0058
%RSD	25.69	340.1	426.6
#1	-.0138	.0116	.0087
#2	-.0153	-.0026	-.0054
#3	-.0143	.0092	-.0022
#4	-.0070	-.0050	-.0052
#5	-.0135	-.0021	-.0027

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.0815
Stddev	.1865
%RSD	2.0533

#1	9.0385
#2	9.1189
#3	8.7845
#4	9.2626
#5	9.2032

Method: I-129 Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/19/99 11:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0535	-.0049	-.0136
Stddev	.0097	.0104	.0080
%RSD	18.21	214.4	58.76

#1	-.0404	-.0034	-.0123
#2	-.0521	.0093	-.0029
#3	-.0500	-.0094	-.0251
#4	-.0664	-.0018	-.0155
#5	-.0588	-.0190	-.0120

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4378
Stddev	.0966
%RSD	1.0232

#1	9.4391
#2	9.4397
#3	9.5282
#4	9.2798
#5	9.5023

3278

Method: I-129

Sample Name: s98t002604

Operator: bjg

Comment: 201 df

Run Time: 03/19/99

11:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I ₁₂₇	I ₁₂₉	Xe ₁₃₁
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0329	-.0090	-.0127
Stddev	.0128	.0107	.0115
%RSD	38.88	117.9	90.95

#1	-.0515	.0084	-.0176
#2	-.0216	-.0090	-.0205
#3	-.0242	-.0183	.0076
#4	-.0407	-.0170	-.0180
#5	-.0263	-.0093	-.0148

Int. Std.	In ₁₁₅
Line	115/pulse
Units	Cts/S
Avg	9.7541
Stddev	.1969
%RSD	2.0191

#1	10.095
#2	9.6843
#3	9.6881
#4	9.7189
#5	9.5847

Method: I-129

Sample Name: s98t002604_d

Operator: bjg

Comment: 201 df

Run Time: 03/19/99 11:55 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I ₋ [127]	I ₋ [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0427	-.0077	-.0116
Stddev	.0042	.0131	.0101
%RSD	9.875	170.1	87.50

#1	-.0426	-.0094	-.0147
#2	-.0429	-.0288	-.0252
#3	-.0484	-.0006	-.0006
#4	-.0365	-.0057	-.0149
#5	-.0432	.0060	-.0024

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.9292
Stddev	.2171
%RSD	2.1867

#1	10.267
#2	9.8644
#3	9.8871
#4	9.9594
#5	9.6686

Method: I-129   Sample Name: s98t002604_a   Operator: bjg
Comment: 201 df
Run Time: 03/19/99 11:57 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	10.11	.0064	.0074
Stddev	.05	.0099	.0181
%RSD	.5154	153.5	245.0

#1	10.13	.0075	-.0048
#2	10.04	.0122	.0005
#3	10.07	.0164	.0127
#4	10.14	-.0096	.0364
#5	10.16	.0058	-.0080

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.9131
Stddev	.0485
%RSD	.48967

#1	9.9679
#2	9.9133
#3	9.8568
#4	9.8734
#5	9.9540

Method: I-129

Sample Name: s98t002605

Operator: bjg

Comment: 201 df

Run Time: 03/19/99 12:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0281	.0075	-.0018
Stddev	.0080	.0024	.0107
%RSD	28.41	32.06	582.1

#1	.0180	.0110	-.0115
#2	.0295	.0086	-.0069
#3	.0338	.0068	.0159
#4	.0371	.0058	-.0061
#5	.0221	.0051	-.0006

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8898
Stddev	.1172
%RSD	1.1849

#1	10.088
#2	9.8944
#3	9.8319
#4	9.7888
#5	9.8457

Method: I-129

Sample Name: s98t002605_d

Operator: bjg

Comment: 201 df

Run Time: 03/19/99 12:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0292	-.0072	-.0088
Stddev	.0135	.0105	.0033
%RSD	46.44	144.4	37.47
#1	-.0381	.0100	-.0116
#2	-.0217	-.0049	-.0097
#3	-.0151	-.0161	-.0118
#4	-.0482	-.0123	-.0069
#5	-.0228	-.0128	-.0041

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.009
Stddev	.107
%RSD	1.0712
#1	10.173
#2	10.003
#3	9.8781
#4	10.021
#5	9.9679

Method: I-129

Sample Name: s98t002606

Operator: bjg

Comment: 201 df

Run Time: 03/19/99 12:06 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0345	-.0062	-.0127
Stddev	.0106	.0158	.0063
%RSD	30.63	254.8	49.82
#1	-.0375	-.0054	-.0160
#2	-.0172	.0006	-.0170
#3	-.0399	-.0117	-.0172
#4	-.0330	.0142	-.0022
#5	-.0448	-.0286	-.0114

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	9.7907
Stddev	.1111
%RSD	1.1348

#1	9.8641
#2	9.7320
#3	9.7980
#4	9.6379
#5	9.9217

Method: I-129   Sample Name: s98t002606_d   Operator: bjg
Comment: 201 df
Run Time: 03/19/99 12:08 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0412	-.0196	-.0204
Stddev	.0126	.0125	.0064
%RSD	30.55	63.94	31.60

#1	-.0624	-.0203	-.0277
#2	-.0300	-.0043	-.0172
#3	-.0416	-.0154	-.0224
#4	-.0345	-.0189	-.0110
#5	-.0375	-.0389	-.0237

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	9.9805
Stddev	.2309
%RSD	2.3138

#1	10.384
#2	9.8162
#3	9.9451
#4	9.8529
#5	9.9041

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/19/99 12:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.009	9.830	.0355
Stddev	.064	.297	.0138
%RSD	3.187	3.018	38.86
#1	1.914	9.387	.0385
#2	2.004	9.749	.0158
#3	2.063	9.870	.0456
#4	1.990	9.953	.0277
#5	2.073	10.19	.0497

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	7.4248
Stddev	.2940
%RSD	3.9601

#1	7.9065
#2	7.5033
#3	7.2569
#4	7.2646
#5	7.1929

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmaH

Run Time: 03/19/99 12:15 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0023	.0013	-.0009
Stddev	.0074	.0093	.0119
%RSD	326.6	704.7	1262.

#1	.0040	.0066	.0089
#2	-.0145	-.0136	-.0166
#3	-.0035	-.0013	-.0049
#4	.0023	.0050	-.0052
#5	.0004	.0099	.0130

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8998
Stddev	.0839
%RSD	.84799

#1	9.8380
#2	10.010
#3	9.9301
#4	9.7965
#5	9.9240

LABCORE Completed Worklist Report for Worklist# 28547

Analyst: bjg

Instrument: ICPMS1

Book#: 990319-3

Method: LA-506-101 Rev/Mod

A-0

Worklist Comment: ICP/MS BY-105 (I129) Solid Acid Digest

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	I129	QC	0.00927	1.01e-02	108.954 % Recovery	
2	ICB	0	I129	QC	1	<1.20e-4	ug/mL	
3	BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3	ug/g	
4	SAMPLE	S98T002713 0 A	I129-A1	SOLID	<u>N/A</u> <	4.786e+00	4.786	ug/g
5	DUP	S98T002713 0 A	I129-A1	SOLID	<4.79e0	<4.69e0	RPD	
6	SAMPLE	S98T002740 0 A	I129-A1	SOLID	<u>N/A</u> <	4.924e+00	4.925	ug/g
7	DUP	S98T002740 0 A	I129-A1	SOLID	<4.92e0	<4.68e0	RPD	
8	SAMPLE	S98T002846 0 A	I129-A1	SOLID	<u>N/A</u> <	4.918e+00	4.918	ug/g
9	DUP	S98T002846 0 A	I129-A1	SOLID	<4.92e0	<4.87e0	RPD	
10	CCV	0	I129	QC	0.00927	9.95e-03	107.335 % Recovery	
11	CCB	0	I129	QC	1	<1.20e-4	ug/mL	
12	BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3	ug/g	
13	SAMPLE	S98T002847 0 A	I129-A1	SOLID	<u>N/A</u> <	5.058e+02	505.900	ug/g
14	DUP	S98T002847 0 A	I129-A1	SOLID	<5.06e2	<4.60e2	RPD	
15	SAMPLE	S98T002926 0 A	I129-A1	SOLID	<u>N/A</u> <	4.776e+00	4.776	ug/g
16	DUP	S98T002926 0 A	I129-A1	SOLID	<4.78e0	<4.83e0	RPD	
17	SAMPLE	S98T002927 0 A	I129-A1	SOLID	<u>N/A</u> <	4.726e+00	4.726	ug/g
18	DUP	S98T002927 0 A	I129-A1	SOLID	<4.73e0	<4.60e0	RPD	
19	CCV	0	I129	QC	0.00927	9.99e-03	107.767 % Recovery	
20	CCB	0	I129	QC	1	<1.20e-4	ug/mL	

Final page for worklist# 28547

Analyst Signature

Date

Analyst Signature

Date

Saul M. Parry
Reviewer Signature

03/23/99
Date

LABCORE Data Entry Template for Worklist# 28547

Analyst: BJL Instrument: ICPMS1 Book# 990318-3Method: LA-506-101 Rev/Mod A-O

Worklist Comment: ICP/MS BY-105 (I129) Solid Acid Digest

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			I129	QC		
2 ICB			I129	QC		
3 BLNK-PREP			I129-A1	SOLID		
4 SAMPLE	S98T002713 0 A		I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
5 DUP	S98T002713 0 A		I129-A1	SOLID		
6 SAMPLE	S98T002740 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
7 DUP	S98T002740 0 A		I129-A1	SOLID		
8 SAMPLE	S98T002846 0 A		I129-A1	SOLID	98000437	BY-105 (2)
	Analytes Requested: I129-A1					
9 DUP	S98T002846 0 A		I129-A1	SOLID		
10 CCV			I129	QC		
11 CCB			I129	QC		
12 BLNK-PREP			I129-A1	SOLID		
13 SAMPLE	S98T002847 0 A		I129-A1	SOLID	98000437	BY-105 (2)
	Analytes Requested: I129-A1					
14 DUP	S98T002847 0 A		I129-A1	SOLID		
15 SAMPLE	S98T002926 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
16 DUP	S98T002926 0 A		I129-A1	SOLID		

Data Entry Comments:

Uploaded & Validated by:
Santana Pang
03/23/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 28547

S Type	Sample#	R A	Test	Matrix	Group#	Project
17 SAMPLE	S98T002927	0 A	I129-A1	SOLID	98000490	BY-105 (2)
Analytes Requested: I129-A1						
18 DUP	S98T002927	0 A	I129-A1	SOLID		
19 CCV			I129	QC		
20 CCB			I129	QC		

Final page for worklist # 28547

3-22-99
Signature Date

Signature Date

PREPBLK
S98T002713
2713-D
2740
2740-D
2846
2846-D

1ml - 9ml
0.05ml - 10ml

PREPBLK
2847
2847-D
2926
2926-D
2927
2927-D

1ml - 9ml
0.05ml - 10ml - 1ml - 10ml
0.05ml - 10ml

D.F.
10
201

10
20,301
201

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/22/99 10:21 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	140300.
Stddev	1558.
%RSD	1.111

#1	142200.
#2	141000.
#3	139800.
#4	139900.
#5	136400.
#6	140000.
#7	141200.
#8	141500.
#9	140300.
#10	140300.

B. Boche
3-22-99

64-105
598700 2713, 2740, 2846-47, 2896-27

FILE A 10322A.TXT
W. L. A 28547

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3291 TO 3316

Method: I-129

Sample Name: Blank

Operator:

Comment:

Run Time: 03/22/99 10:23 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	70.71	144.5	118.3
Stddev	.97	3.4	4.0
%RSD	1.378	2.361	3.408

#1	71.77	149.3	124.0
#2	70.70	145.8	117.6
#3	69.58	140.0	114.1
#4	71.57	143.5	115.2
#5	69.91	143.8	120.5

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.703
Stddev	.136
%RSD	1.2739

#1	10.533
#2	10.587
#3	10.843
#4	10.807
#5	10.743

HNF-1663 REV. 0

Method : I-129

Standardzn Report

03/22/99 10:24:47

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.1904	70.7074	1.0000000	03/22/99 10:24:30
I_[129]	203.7753	144.4827	1.0000000	03/22/99 10:24:30
Xe[131]	205.3432	118.2832	1.0000000	03/22/99 10:24:30

3293

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/22/99 10:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_ [127]	I_ [129]	Xe [131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3806.	3806.	3806.
Stddev	50.	50.	50.
%RSD	1.317	1.317	1.317
#1	3727.	3727.	3727.
#2	3791.	3791.	3791.
#3	3819.	3819.	3819.
#4	3848.	3848.	3848.
#5	3847.	3847.	3847.

Int. Std.	In [115]
Line	115/pulse
Units	Cts/S
Avg	10.912
Stddev	.096
%RSD	.88085

#1	11.030
#2	10.944
#3	10.922
#4	10.899
#5	10.764

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	223.5636	70.7074	1.0000000	03/22/99 10:26:42
I_[129]	219.1486	144.4827	1.0000000	03/22/99 10:26:42
Xe[131]	220.7165	118.2832	1.0000000	03/22/99 10:26:42

Method: I-129 Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/22/99 10:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0029	-.0012	-.0058
Stddev	.0080	.0076	.0132
%RSD	274.6	635.4	227.4

#1	.0096	.0017	.0103
#2	-.0052	-.0037	-.0108
#3	-.0056	.0051	-.0081
#4	.0115	.0042	.0033
#5	.0043	-.0133	-.0237

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.109
Stddev	.080
%RSD	.71806

#1	11.181
#2	11.178
#3	11.129
#4	11.062
#5	10.995

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/22/99 10:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.999	10.06	.0465
Stddev	.049	.33	.0147
%RSD	2.448	3.291	31.71
#1	2.053	10.54	.0661
#2	1.934	9.930	.0465
#3	1.962	9.633	.0330
#4	2.029	10.03	.0552
#5	2.015	10.14	.0315

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.3301
Stddev	.1762
%RSD	2.1149

#1	8.0639
#2	8.4749
#3	8.5056
#4	8.3092
#5	8.2969

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 10:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0127	-.0358	-.0165
Stddev	.0091	.0153	.0096
%RSD	71.85	42.75	58.03

#1	-.0083	-.0220	-.0038
#2	-.0056	-.0295	-.0131
#3	-.0080	-.0487	-.0222
#4	-.0134	-.0234	-.0144
#5	-.0283	-.0555	-.0291

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.368
Stddev	.087
%RSD	.76526

#1	11.410
#2	11.383
#3	11.236
#4	11.342
#5	11.469

3298

Method: I-129   Sample Name: prepblk   Operator: bjg
Comment: 10 df
Run Time: 03/22/99 10:51 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1530	.0032	-.0019
Stddev	.0023	.0178	.0031
%RSD	1.491	550.6	162.4

#1	-.1558	-.0006	.0000
#2	-.1537	.0096	-.0045
#3	-.1496	-.0257	.0019
#4	-.1537	.0125	-.0013
#5	-.1520	.0204	-.0055

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.387
Stddev	.131
%RSD	1.1478

#1	11.523
#2	11.322
#3	11.530
#4	11.310
#5	11.249

Method: I-129

Sample Name: s98t002713

Operator: bjg

Comment: 201 df

Run Time: 03/22/99

10:53 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0886	.0090	-.0058
Stddev	.0221	.0190	.0062
%RSD	24.95	212.5	107.7

#1	.1264	.0356	-.0047
#2	.0861	.0052	-.0030
#3	.0824	.0018	-.0130
#4	.0796	-.0155	.0026
#5	.0686	.0177	-.0108

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.536
Stddev	.045
%RSD	.39423

#1	11.489
#2	11.516
#3	11.606
#4	11.515
#5	11.552

Method: I-129

Sample Name: s98t002713_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 10:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0604	-.0111	-.0102
Stddev	.0059	.0161	.0098
%RSD	9.773	144.3	95.75
#1	.0656	.0000	-.0085
#2	.0678	-.0217	-.0109
#3	.0567	-.0023	-.0187
#4	.0548	-.0342	-.0184
#5	.0568	.0025	.0053

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.748
Stddev	.088
%RSD	.75114

#1	11.683
#2	11.676
#3	11.865
#4	11.819
#5	11.694

Method: I-129   Sample Name: s98t002740   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 10:56   Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0330	.0025	.0052
Stddev	.0090	.0120	.0063
%RSD	27.15	474.7	120.4

#1	.0345	.0124	.0071
#2	.0262	-.0071	-.0054
#3	.0443	.0078	.0105
#4	.0220	.0128	.0053
#5	.0379	-.0133	.0085

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.871
Stddev	.126
%RSD	1.0619

#1	12.007
#2	12.009
#3	11.803
#4	11.787
#5	11.751

Method: I-129

Sample Name: s98t002740_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 10:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0355	-.0046	-.0072
Stddev	.0086	.0112	.0091
%RSD	24.24	245.0	125.7

#1	.0432	.0046	-.0203
#2	.0455	-.0131	-.0069
#3	.0257	-.0063	.0054
#4	.0339	-.0172	-.0062
#5	.0293	.0091	-.0081

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.902
Stddev	.027
%RSD	.22305

#1	11.884
#2	11.875
#3	11.928
#4	11.932
#5	11.889

3303

HNF-1663 REV. 0

2846

Method: I-129

Sample Name: s98t002746 ~~1632258~~

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 10:59 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0459	-.0094	.0058
Stddev	.0087	.0065	.0072
%RSD	18.90	69.31	124.6

#1	.0474	-.0098	-.0031
#2	.0364	-.0047	.0129
#3	.0375	-.0023	.0041
#4	.0533	-.0111	.0135
#5	.0551	-.0192	.0016

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.939
Stddev	.148
%RSD	1.2363

#1	12.126
#2	11.872
#3	12.068
#4	11.826
#5	11.805

3304

Method: I-129

Sample Name: s98t002746-d
2846-d
9/6/99

Operator: bjg

Comment: 201 df

Run Time: 03/22/99

11:01 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0340	-.0073	-.0072
Stddev	.0090	.0077	.0124
%RSD	26.43	104.7	171.6

#1	.0220	.0002	-.0169
#2	.0355	-.0020	-.0030
#3	.0307	-.0196	-.0064
#4	.0348	-.0078	-.0204
#5	.0468	-.0076	.0108

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.006
Stddev	.100
%RSD	.83310

#1	11.987
#2	12.010
#3	12.112
#4	12.071
#5	11.851

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 11:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.009	9.951	.0412
Stddev	.028	.184	.0210
%RSD	1.403	1.849	.51.08
#1	1.967	9.710	.0110
#2	2.025	9.904	.0293
#3	2.013	9.871	.0465
#4	2.041	10.16	.0563
#5	1.998	10.11	.0627

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.6039
Stddev	.2493
%RSD	2.5957

#1	9.9817
#2	9.7033
#3	9.4912
#4	9.3297
#5	9.5139

Method: I-129   Sample Name: ccb   Operator: bjg
Comment: 2% tmah
Run Time: 03/22/99 11:06 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0070	-.0231	-.0086
Stddev	.0037	.0012	.0151
%RSD	53.01	5.346	176.3

#1	-.0126	-.0225	-.0345
#2	-.0079	-.0252	-.0051
#3	-.0054	-.0221	-.0067
#4	-.0027	-.0230	.0038
#5	-.0062	-.0228	-.0003

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.379
Stddev	.099
%RSD	.79970

#1	12.389
#2	12.389
#3	12.525
#4	12.252
#5	12.342

Method: I-129   Sample Name: prepblk   Operator: bjg
Comment: 10 df
Run Time: 03/22/99 11:08 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_ [127]	I_ [129]	Xe [131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1579	-.0115	-.0117
Stddev	.0104	.0216	.0203
%RSD	6.609	188.0	173.4

#1	-.1729	-.0499	-.0458
#2	-.1506	.0018	-.0057
#3	-.1639	-.0061	-.0095
#4	-.1544	-.0014	.0015
#5	-.1475	-.0019	-.0103

Int. Std.	In [115]
Line	115/pulse
Units	Cts/S
Avg	11.871
Stddev	.503
%RSD	4.2332

#1	12.748
#2	11.486
#3	11.800
#4	11.635
#5	11.689

3308

Method: I-129   Sample Name: s98t002847   Operator: bjg
Comment: 20301 df
Run Time: 03/22/99 11:10 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe [131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0554	-.0064	-.0086
Stddev	.0184	.0095	.0103
%RSD	33.22	147.7	120.3

#1	.0824	-.0145	.0059
#2	.0648	-.0125	-.0160
#3	.0473	-.0128	-.0042
#4	.0352	.0038	-.0204
#5	.0475	.0041	-.0080

Int. Std.	In [115]
Line	115/pulse
Units	Cts/S
Avg	12.142
Stddev	.069
%RSD	.57138

#1	12.230
#2	12.166
#3	12.169
#4	12.054
#5	12.093

Method: I-129

Sample Name: s98t002847_d

Operator: bjg

Comment: 20301 df

Run Time: 03/22/99

11:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0321	-.0127	.0005
Stddev	.0096	.0129	.0179
%RSD	30.01	101.6	3481.

#1	.0327	-.0186	-.0152
#2	.0455	.0048	.0023
#3	.0332	-.0145	.0288
#4	.0184	-.0060	-.0149
#5	.0305	-.0293	.0016

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.211
Stddev	.085
%RSD	.69825

#1	12.304
#2	12.185
#3	12.147
#4	12.122
#5	12.299

Method: I-129

Sample Name: s98t002926

Operator: bjg

Comment: 201 df

Run Time: 03/22/99

11:13 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0285	-.0064	.0044
Stddev	.0076	.0064	.0120
%RSD	26.76	100.3	269.6
#1	.0322	.0032	-.0004
#2	.0231	-.0053	.0092
#3	.0355	-.0098	.0192
#4	.0339	-.0061	-.0130
#5	.0179	-.0141	.0073

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.293
Stddev	.084
%RSD	.68732
#1	12.404
#2	12.318
#3	12.182
#4	12.242
#5	12.320

Method: I-129   Sample Name: s98t002926_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 11:14   Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0279	-.0095	-.0030
Stddev	.0065	.0087	.0100
%RSD	23.42	91.36	330.6

#1	.0350	-.0110	-.0051
#2	.0316	-.0095	.0006
#3	.0250	.0039	-.0078
#4	.0297	-.0203	.0119
#5	.0182	-.0105	-.0148

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	12.311
Stddev	.018
%RSD	.14954

#1	12.287
#2	12.338
#3	12.306
#4	12.315
#5	12.312

3313

Method: I-129   Sample Name: s98t002927   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 11:16 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0304	-.0044	.0091
Stddev	.0119	.0191	.0129
%RSD	39.18	438.3	140.8

#1	.0148	-.0016	-.0004
#2	.0372	-.0296	.0259
#3	.0230	.0027	.0006
#4	.0315	.0214	-.0006
#5	.0453	-.0147	.0202

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.109
Stddev	.107
%RSD	.88764

#1	12.281
#2	11.996
#3	12.055
#4	12.128
#5	12.087

Method: I-129   Sample Name: s98t002927_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 11:18 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0221	-.0090	-.0134
Stddev	.0126	.0056	.0159
%RSD	56.83	62.52	119.3

#1	.0131	-.0139	-.0117
#2	.0088	-.0129	-.0359
#3	.0376	.0002	-.0111
#4	.0330	-.0098	-.0087
#5	.0180	-.0084	-.0167

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.240
Stddev	.126
%RSD	1.0287

#1	12.316
#2	12.396
#3	12.174
#4	12.071
#5	12.244

HNF-1663 REV. 0

Method: I-129

Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 11:20 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.988	9.994	.0497
Stddev	.081	.222	.0121
%RSD	4.079	2.219	24.34
#1	1.856	9.618	.0411
#2	1.969	9.998	.0641
#3	2.030	10.04	.0563
#4	2.062	10.17	.0339
#5	2.022	10.14	.0531

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.7659
Stddev	.3948
%RSD	4.0426

#1	10.408
#2	9.8434
#3	9.6365
#4	9.5578
#5	9.3842

Method: I-129

Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99

11:22 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0095	-.0224	-.0146
Stddev	.0086	.0080	.0107
%RSD	90.40	35.92	73.64

#1	-.0116	-.0207	-.0156
#2	-.0093	-.0320	-.0257
#3	.0050	-.0173	.0031
#4	-.0146	-.0290	-.0151
#5	-.0169	-.0127	-.0197

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.517
Stddev	.078
%RSD	.61976

#1	12.561
#2	12.573
#3	12.386
#4	12.558
#5	12.506

LABCORE Completed Worklist Report for Worklist# 28557

Analyst: bjg

Instrument: ICPMS1

Book#: 990319-3

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	I129	QC	0.00927	1.01e-02	108.954	% Recovery
2	ICB	0	I129	QC	1	<1.20e-4		ug/mL
3	BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3		ug/g
4	SAMPLE	S98T002928	0 A	I129-A1	SOLID	N/A <	4.918e+00	4.918 ug/g
5	DUP	S98T002928	0 A	I129-A1	SOLID	<4.92e0	<4.58e0	RPD
6	SAMPLE	S98T002929	0 A	I129-A1	SOLID	N/A <	4.744e+00	4.744 ug/g
7	DUP	S98T002929	0 A	I129-A1	SOLID	<4.74e0	<4.90e0	RPD
8	SAMPLE	S98T002930	0 A	I129-A1	SOLID	N/A <	4.736e+02	473.600 ug/g
9	DUP	S98T002930	0 A	I129-A1	SOLID	<4.74e2	<4.81e2	RPD
10	BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3		ug/g
11	SAMPLE	S98T002990	0 A	I129-A1	SOLID	N/A <	4.678e+00	4.677 ug/g
12	DUP	S98T002990	0 A	I129-A1	SOLID	<4.68e0	<4.80e0	RPD
13	CCV	0	I129	QC	0.00927	9.73e-03	104.962	% Recovery
14	CCB	0	I129	QC	1	<1.20e-4		ug/mL
15	SAMPLE	S98T002991	0 A	I129-A1	SOLID	N/A <	4.782e+00	4.782 ug/g
16	DUP	S98T002991	0 A	I129-A1	SOLID	<4.78e0	<4.90e0	RPD
17	SAMPLE	S98T002992	0 A	I129-A1	SOLID	N/A <	4.713e+00	4.713 ug/g
18	DUP	S98T002992	0 A	I129-A1	SOLID	<4.71e0	<4.78e0	RPD
19	SAMPLE	S98T002993	0 A	I129-A1	SOLID	N/A <	4.811e+00	4.810 ug/g
20	DUP	S98T002993	0 A	I129-A1	SOLID	<4.81e0	<4.95e0	RPD
21	CCV	0	I129	QC	0.00927	1.02e-02	110.032	% Recovery
22	CCB	0	I129	QC	1	<1.20e-4		ug/mL

Final page for worklist# 28557

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul M. Pang 03/23/99
 Reviewer Signature _____ Date _____

FILE # 10322A.705

03/10/99 08:40

ws2

Page: 1

LABCORE Data Entry Template for Worklist# 28557

Analyst: SL Instrument: ICPMS1 _____ Book# 990379-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			I129	QC		
2 ICB			I129	QC		
3 BLNK-PREP			I129-A1	SOLID		
4 SAMPLE	S98T002928 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
5 DUP	S98T002928 0 A		I129-A1	SOLID		
6 SAMPLE	S98T002929 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
7 DUP	S98T002929 0 A		I129-A1	SOLID		
8 SAMPLE	S98T002930 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
9 DUP	S98T002930 0 A		I129-A1	SOLID		
10 BLNK-PREP			I129-A1	SOLID		
11 SAMPLE	S98T002990 0 A		I129-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: I129-A1					
12 DUP	S98T002990 0 A		I129-A1	SOLID		
13 CCV			I129	QC		
14 CCB			I129	QC		
15 SAMPLE	S98T002991 0 A		I129-A1	SOLID	98000509	BY-105 (2)
	Analytes Requested: I129-A1					
16 DUP	S98T002991 0 A		I129-A1	SOLID		

Data Entry Comments:

Uploaded + Validated by
 Saul M. Paring
 03/23/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3318

LABCORE Data Entry Template for Worklist# 28557

S Type	Sample#	R	A	Test	Matrix	Group#	Project
17 SAMPLE	S98T002992	0	A	I129-A1	SOLID	98000509	BY-105 (2)
Analytes Requested: I129-A1							
18 DUP	S98T002992	0	A	I129-A1	SOLID		
19 SAMPLE	S98T002993	0	A	I129-A1	SOLID	98000509	BY-105 (2)
Analytes Requested: I129-A1							
20 DUP	S98T002993	0	A	I129-A1	SOLID		
21 CCV				I129	QC		
22 CCB				I129	QC		

Final page for worklist # 28557

[Signature] 3-22-99
Signature Date

Signature Date

PREP BLK
S98T00 2994
2994-D
2999
2999-D
2990
2990-D

1ml-8ml
0.05ml-10ml
↓
0.05ml-10ml 0.1ml-10ml
↓ ↓

DF
10
201
↓
20,301
↓

PREP BLK
S98T00 2990
2990-D
2991
2991-D
2992
2992-D
2993
2993-D

1ml-8ml
0.05ml-10ml
↓

10
201
↓

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3319

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/22/99

10:21 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	140300.
Stddev	1558.
%RSD	1.111

#1	142200.
#2	141000.
#3	139800.
#4	139900.
#5	136400.
#6	140000.
#7	141200.
#8	141500.
#9	140300.
#10	140300.

8
3/22/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3320 TO 3347

Method: I-129

Sample Name: Blank

Operator:

Comment:

Run Time: 03/22/99 10:23 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	70.71	144.5	118.3
Stddev	.97	3.4	4.0
%RSD	1.378	2.361	3.408
#1	71.77	149.3	124.0
#2	70.70	145.8	117.6
#3	69.58	140.0	114.1
#4	71.57	143.5	115.2
#5	69.91	143.8	120.5

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.703
Stddev	.136
%RSD	1.2739

#1	10.533
#2	10.587
#3	10.843
#4	10.807
#5	10.743

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.1904	70.7074	1.0000000	03/22/99 10:24:30
I_[129]	203.7753	144.4827	1.0000000	03/22/99 10:24:30
Xe[131]	205.3432	118.2832	1.0000000	03/22/99 10:24:30

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/22/99 10:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3806.	3806.	3806.
Stddev	50.	50.	50.
%RSD	1.317	1.317	1.317
#1	3727.	3727.	3727.
#2	3791.	3791.	3791.
#3	3819.	3819.	3819.
#4	3848.	3848.	3848.
#5	3847.	3847.	3847.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.912
Stddev	.096
%RSD	.88085

#1	11.030
#2	10.944
#3	10.922
#4	10.899
#5	10.764

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	223.5636	70.7074	1.0000000	03/22/99 10:26:42
I_[129]	219.1486	144.4827	1.0000000	03/22/99 10:26:42
Xe[131]	220.7165	118.2832	1.0000000	03/22/99 10:26:42

Method: I-129 Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/22/99 10:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0029	-.0012	-.0058
Stddev	.0080	.0076	.0132
%RSD	274.6	635.4	227.4

#1	.0096	.0017	.0103
#2	-.0052	-.0037	-.0108
#3	-.0056	.0051	-.0081
#4	.0115	.0042	.0033
#5	.0043	-.0133	-.0237

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.109
Stddev	.080
%RSD	.71806

#1	11.181
#2	11.178
#3	11.129
#4	11.062
#5	10.995

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/22/99 10:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.999	10.06	.0465
Stddev	.049	.33	.0147
%RSD	2.448	3.291	31.71
#1	2.053	10.54	.0661
#2	1.934	9.930	.0465
#3	1.962	9.633	.0330
#4	2.029	10.03	.0552
#5	2.015	10.14	.0315

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	8.3301
Stddev	.1762
%RSD	2.1149

#1	8.0639
#2	8.4749
#3	8.5056
#4	8.3092
#5	8.2969

HNF-1663 REV. 0

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 10:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I ₋ [127]	I ₋ [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0127	-.0358	-.0165
Stddev	.0091	.0153	.0096
%RSD	71.85	42.75	58.03
#1	-.0083	-.0220	-.0038
#2	-.0056	-.0295	-.0131
#3	-.0080	-.0487	-.0222
#4	-.0134	-.0234	-.0144
#5	-.0283	-.0555	-.0291

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.368
Stddev	.087
%RSD	.76526

#1	11.410
#2	11.383
#3	11.236
#4	11.342
#5	11.469

Method: I-129 Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/22/99 11:40 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1568	-.0409	-.0265
Stddev	.0092	.0257	.0166
%RSD	5.890	62.94	62.43
#1	-.1682	-.0661	-.0518
#2	-.1617	-.0703	-.0157
#3	-.1582	-.0320	-.0339
#4	-.1518	-.0153	-.0206
#5	-.1442	-.0207	-.0107

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.490
Stddev	.454
%RSD	3.6325

#1	13.265
#2	12.428
#3	12.365
#4	12.322
#5	12.072

Method: I-129 Sample Name: s98t002928

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 11:42 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0443	.0030	.0202
Stddev	.0245	.0111	.0154
%RSD	55.26	364.4	76.27

#1	.0860	.0170	.0417
#2	.0395	.0107	.0306
#3	.0422	.0029	.0100
#4	.0229	-.0091	.0137
#5	.0311	-.0064	.0049

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	12.053
Stddev	.189
%RSD	1.5675

#1	11.848
#2	11.844
#3	12.203
#4	12.179
#5	12.189

Method: I-129

Sample Name: s98t002928_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 11:43 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0125	-.0179	-.0154
Stddev	.0092	.0259	.0161
%RSD	73.13	144.2	104.3
#1	.0197	-.0455	-.0401
#2	.0148	-.0005	.0047
#3	.0007	-.0305	-.0111
#4	.0220	.0178	-.0148
#5	.0055	-.0310	-.0157

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.429
Stddev	.145
%RSD	1.1675

#1	12.486
#2	12.308
#3	12.475
#4	12.258
#5	12.616

Method: I-129

Sample Name: s98t002929

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 11:45 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0104	-.0146	-.0050
Stddev	.0074	.0131	.0128
%RSD	71.34	89.51	258.1
#1	.0135	-.0161	-.0254
#2	.0212	.0012	-.0014
#3	.0092	-.0198	-.0083
#4	.0063	-.0327	.0033
#5	.0017	-.0057	.0070

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	12.427
Stddev	.145
%RSD	1.1651

#1	12.567
#2	12.194
#3	12.417
#4	12.431
#5	12.525

Method: I-129

Sample Name: s98t002929_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99

11:46 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0198	-.0058	-.0069
Stddev	.0128	.0131	.0112
%RSD	64.89	225.0	162.5

#1	.0192	.0071	-.0054
#2	.0330	.0016	-.0057
#3	.0117	-.0188	-.0234
#4	.0317	.0020	-.0081
#5	.0032	-.0209	.0081

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.565
Stddev	.206
%RSD	1.6403

#1	12.400
#2	12.729
#3	12.718
#4	12.289
#5	12.692

Method: I-129

Sample Name: s98t002930

Operator: bjg

Comment: 20301 df

Run Time: 03/22/99

11:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0100	-.0181	-.0083
Stddev	.0105	.0189	.0125
%RSD	105.5	104.3	151.7

#1	-.0031	.0093	.0029
#2	.0116	-.0308	-.0119
#3	.0251	-.0112	-.0156
#4	.0122	-.0396	.0064
#5	.0041	-.0181	-.0231

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.783
Stddev	.225
%RSD	1.7599

#1	13.062
#2	12.655
#3	12.604
#4	12.603
#5	12.991

3333

Method: I-129   Sample Name: s98t002930_d   Operator: bjg
Comment: 20301 df
Run Time: 03/22/99 11:50 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0064	-.0184	-.0174
Stddev	.0103	.0179	.0115
%RSD	160.0	97.50	66.02

#1	-.0051	-.0090	-.0215
#2	.0230	-.0315	-.0076
#3	.0067	-.0113	-.0356
#4	.0051	.0019	-.0131
#5	.0025	-.0422	-.0093

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.818
Stddev	.124
%RSD	.96986

#1	12.921
#2	12.966
#3	12.810
#4	12.688
#5	12.707

Method: I-129 Sample Name: prepbkl

Operator: bjg

Comment: 10 df

Run Time: 03/22/99 11:54 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I ₋ [127]	I ₋ [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1620	-.0336	-.0335
Stddev	.0096	.0392	.0441
%RSD	5.904	116.6	131.6

#1	-.1774	-.1022	-.1105
#2	-.1651	-.0218	-.0243
#3	-.1570	-.0096	-.0228
#4	-.1534	-.0078	-.0096
#5	-.1572	-.0265	-.0005

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.639
Stddev	1.123
%RSD	8.8863

#1	14.602
#2	12.512
#3	12.204
#4	11.894
#5	11.981

3335

Method: I-129   Sample Name: s98t002990   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 11:56 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0179	-.0227	-.0159
Stddev	.0274	.0220	.0102
%RSD	152.9	96.69	64.64

#1	-.0292	-.0182	-.0330
#2	.0378	-.0446	-.0070
#3	.0230	-.0085	-.0108
#4	.0368	.0036	-.0113
#5	.0213	-.0459	-.0173

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.847
Stddev	1.184
%RSD	9.2176

#1	14.936
#2	12.482
#3	12.247
#4	12.043
#5	12.527

Method: I-129

Sample Name: s98t002990_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99

11:57 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0184	-.0143	-.0101
Stddev	.0130	.0160	.0115
%RSD	70.98	111.8	113.7
#1	.0234	-.0377	-.0182
#2	.0352	-.0110	-.0044
#3	.0102	-.0076	-.0095
#4	.0218	-.0207	.0054
#5	.0013	.0053	-.0241

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.347
Stddev	.111
%RSD	.89790

#1	12.483
#2	12.318
#3	12.338
#4	12.187
#5	12.410

3337

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 12:00 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.980	9.726	.0224
Stddev	.180	.659	.0172
%RSD	9.110	6.780	76.68
#1	1.667	8.548	-.0017
#2	2.063	9.960	.0292
#3	2.117	10.02	.0442
#4	2.059	10.05	.0263
#5	1.995	10.06	.0142

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.221
Stddev	.780
%RSD	7.6263
#1	11.611
#2	9.8934
#3	9.9347
#4	9.9006
#5	9.7671

3338

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 12:02 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0025	-.0369	-.0201
Stddev	.0049	.0165	.0097
%RSD	196.0	44.86	48.24
#1	.0093	-.0367	-.0266
#2	-.0024	-.0427	-.0123
#3	.0056	-.0385	-.0336
#4	.0018	-.0105	-.0162
#5	-.0017	-.0560	-.0116

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.985
Stddev	.166
%RSD	1.2757

#1	13.051
#2	13.031
#3	13.057
#4	12.692
#5	13.096

3339

Method: I-129

Sample Name: s98t002991

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 12:04 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0005	-.0375	-.0308
Stddev	.0075	.0085	.0087
%RSD	1560.	22.57	28.14

#1	.0016	-.0501	-.0425
#2	-.0022	-.0336	-.0270
#3	-.0071	-.0404	-.0261
#4	.0114	-.0274	-.0213
#5	-.0061	-.0361	-.0369

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.138
Stddev	.113
%RSD	.85739

#1	13.150
#2	13.146
#3	13.287
#4	12.969
#5	13.137

3340

Method: I-129   Sample Name: s98t002991_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 12:06 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0027	-.0149	-.0121
Stddev	.0050	.0145	.0102
%RSD	185.7	97.33	83.97

#1	-.0017	.0002	-.0240
#2	-.0077	-.0085	-.0003
#3	.0051	-.0116	-.0035
#4	-.0065	-.0385	-.0131
#5	-.0028	-.0161	-.0198

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	12.872
Stddev	.215
%RSD	1.6692

#1	12.945
#2	12.740
#3	12.568
#4	13.017
#5	13.092

Method: I-129   Sample Name: s98t002992   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 12:07 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0049	-.0304	-.0126
Stddev	.0184	.0168	.0163
%RSD	375.1	55.37	129.7

#1	-.0378	-.0385	-.0375
#2	.0044	-.0459	-.0011
#3	.0025	-.0432	-.0130
#4	.0039	-.0137	.0047
#5	.0025	-.0107	-.0160

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	13.206
Stddev	.502
%RSD	3.7983

#1	14.062
#2	13.071
#3	13.147
#4	12.748
#5	13.002

Method: I-129   Sample Name: s98t002992_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 12:09 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_ _[127]	I_ _[129]	Xe _[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0001	-.0278	-.0157
Stddev	.0058	.0107	.0089
%RSD	7658.	38.60	56.52

#1	-.0054	-.0335	-.0098
#2	-.0007	-.0296	-.0306
#3	.0021	-.0292	-.0082
#4	.0089	-.0373	-.0145
#5	-.0045	-.0095	-.0153

Int. Std.	In _[115]
Line	115/pulse
Units	Cts/S
Avg	13.028
Stddev	.119
%RSD	.91689

#1	13.181
#2	13.096
#3	13.046
#4	12.904
#5	12.912

Method: I-129 Sample Name: s98t002993

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 12:10 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0043	-.0150	-.0165
Stddev	.0099	.0228	.0124
%RSD	231.8	151.2	75.32

#1	-.0212	-.0419	-.0311
#2	.0041	-.0101	-.0066
#3	.0004	-.0257	-.0240
#4	-.0034	-.0173	-.0198
#5	-.0011	.0197	-.0010

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.919
Stddev	.217
%RSD	1.6827

#1	13.203
#2	12.698
#3	13.094
#4	12.818
#5	12.782

Method: I-129   Sample Name: s98t002993_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 12:12 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0026	-.0209	-.0176
Stddev	.0084	.0097	.0101
%RSD	323.4	46.51	57.68

#1	-.0122	-.0137	-.0220
#2	-.0097	-.0230	-.0020
#3	.0031	-.0313	-.0249
#4	.0079	-.0083	-.0129
#5	-.0021	-.0283	-.0261

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.913
Stddev	.085
%RSD	.65549

#1	12.811
#2	12.898
#3	12.866
#4	12.964
#5	13.028

Method: I-129

Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 12:14 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.021	10.15	.0330
Stddev	.101	.59	.0200
%RSD	4.982	5.831	60.63
#1	1.853	9.161	.0077
#2	2.011	10.34	.0304
#3	2.053	10.07	.0231
#4	2.106	10.62	.0602
#5	2.084	10.55	.0436

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.7446
Stddev	.4867
%RSD	4.9940

#1	10.569
#2	9.7004
#3	9.6853
#4	9.3517
#5	9.4164

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 12:17 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0061	-.0100	-.0190
Stddev	.0090	.0060	.0089
%RSD	147.7	60.23	47.01
#1	.0021	-.0114	-.0235
#2	.0005	-.0012	-.0179
#3	-.0030	-.0156	-.0149
#4	.0123	-.0070	-.0311
#5	.0186	-.0150	-.0075

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.791
Stddev	.125
%RSD	.98046

#1	12.861
#2	12.813
#3	12.909
#4	12.790
#5	12.582

3347

LABCORE Completed Worklist Report for Worklist# 28558

Analyst: bijg

Instrument: ICPMS1

Book#: 990319-3

Method: LA-506-101 Rev/Mod A-D

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	I129	QC	0.00927	1.01e-02	108.954 % Recovery	
2	ICB	0	I129	QC	1	<1.20e-4	ug/mL	
3	BLAK-PREP	0	I129-A1	SOLID	1	<1.20e-3	ug/g	
4	SAMPLE	S98T003389	0 A	I129-A1	SOLID	<u>N/A</u> <	4.945e+00	4.945 ug/g
5	DUP	S98T003389	0 A	I129-A1	SOLID	<4.95e0	<4.58e0	RPD
6	SAMPLE	S98T003394	0 A	I129-A1	SOLID	<u>N/A</u> <	4.741e+00	4.740 ug/g
7	DUP	S98T003394	0 A	I129-A1	SOLID	<4.74e0	<4.96e0	RPD
8	SAMPLE	S98T003399	0 A	I129-A1	SOLID	<u>N/A</u> <	4.886e+02	488.600 ug/g
9	DUP	S98T003399	0 A	I129-A1	SOLID	<4.89e2	<4.67e2	RPD
10	CCV	0	I129	QC	0.00927	1.01e-02	108.954 % Recovery	
11	CCB	0	I129	QC	1	<1.20e-4	ug/mL	

Final page for worklist# 28558

Analyst Signature

Date

Analyst Signature

Date

Sand H. Pang
Reviewer Signature

03/23/99
Date

03/10/99 08:43
ws2

FILED 10312A-THT 3-22-88
10322A-THT

HNF-1663 REV. 0

Page: 1

LABCORE Data Entry Template for Worklist# 28558

Analyst: STC Instrument: ICPMS1 Book# 990318-3

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (SOLID ACID DIGEST)

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 ICV			I129	QC		
2 ICB			I129	QC		
3 BLNK-PREP			I129-A1	SOLID		
4 SAMPLE	S98T003389 0 A		I129-A1	SOLID	98000490	BY-105 (2)
	Analytes Requested: I129-A1					
5 DUP	S98T003389 0 A		I129-A1	SOLID		
6 SAMPLE	S98T003394 0 A		I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
7 DUP	S98T003394 0 A		I129-A1	SOLID		
8 SAMPLE	S98T003399 0 A		I129-A1	SOLID	98000436	BY-105 (2)
	Analytes Requested: I129-A1					
9 DUP	S98T003399 0 A		I129-A1	SOLID		
10 CCV			I129	QC		
11 CCB			I129	QC		

Final page for worklist # 28558

Signature: [Signature] Date: 3-22-88
3-18-88 3-22-88
JRC/BLK
5557003389
3389-D
3389
3389-D
3389
3389-D
1-1-9-1
05-10-1
1-1-10-1-05-1-10-1
10
10
201
20,301

Data Entry Comments:

Signature Date

Uploaded + Validated by:

Saul M. Pany

03/20/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3349

Method: MS_TUNE   Sample Name: ms-tune   Operator: bjg
Comment:
Run Time: 03/22/99 10:21 Type: Unk   Mode: INT   Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	140300.
Stddev	1558.
%RSD	1.111

#1	142200.
#2	141000.
#3	139800.
#4	139900.
#5	136400.
#6	140000.
#7	141200.
#8	141500.
#9	140300.
#10	140300.

SP 3/22/99

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3350 TO 3366

Method: I-129 Sample Name: Blank

Operator:

Comment:

Run Time: 03/22/99 10:23 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	70.71	144.5	118.3
Stddev	.97	3.4	4.0
%RSD	1.378	2.361	3.408

#1	71.77	149.3	124.0
#2	70.70	145.8	117.6
#3	69.58	140.0	114.1
#4	71.57	143.5	115.2
#5	69.91	143.8	120.5

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.703
Stddev	.136
%RSD	1.2739

#1	10.533
#2	10.587
#3	10.843
#4	10.807
#5	10.743

Method : I-129

Standardzn Report

03/22/99 10:24:47

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.1904	70.7074	1.0000000	03/22/99 10:24:30
I_[129]	203.7753	144.4827	1.0000000	03/22/99 10:24:30
Xe[131]	205.3432	118.2832	1.0000000	03/22/99 10:24:30

3352

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/22/99 10:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3806.	3806.	3806.
Stddev	50.	50.	50.
%RSD	1.317	1.317	1.317
#1	3727.	3727.	3727.
#2	3791.	3791.	3791.
#3	3819.	3819.	3819.
#4	3848.	3848.	3848.
#5	3847.	3847.	3847.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.912
Stddev	.096
%RSD	.88085

#1	11.030
#2	10.944
#3	10.922
#4	10.899
#5	10.764

Method : I-129

Standardzn Report

03/22/99 10:26:49

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	223.5636	70.7074	1.0000000	03/22/99 10:26:42
I_[129]	219.1486	144.4827	1.0000000	03/22/99 10:26:42
Xe[131]	220.7165	118.2832	1.0000000	03/22/99 10:26:42

3354

Method: I-129 Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/22/99 10:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0029	-.0012	-.0058
Stddev	.0080	.0076	.0132
%RSD	274.6	635.4	227.4

#1	.0096	.0017	.0103
#2	-.0052	-.0037	-.0108
#3	-.0056	.0051	-.0081
#4	.0115	.0042	.0033
#5	.0043	-.0133	-.0237

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.109
Stddev	.080
%RSD	.71806

#1	11.181
#2	11.178
#3	11.129
#4	11.062
#5	10.995

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/22/99 10:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.999	10.06	.0465
Stddev	.049	.33	.0147
%RSD	2.448	3.291	31.71

#1	2.053	10.54	.0661
#2	1.934	9.930	.0465
#3	1.962	9.633	.0330
#4	2.029	10.03	.0552
#5	2.015	10.14	.0315

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	8.3301
Stddev	.1762
%RSD	2.1149

#1	8.0639
#2	8.4749
#3	8.5056
#4	8.3092
#5	8.2969

Method: I-129   Sample Name: ich   Operator: bjg
Comment: 2% tmah
Run Time: 03/22/99 10:33 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_ ₁₂₇	I_ ₁₂₉	Xe ₁₃₁
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0127	-.0358	-.0165
Stddev	.0091	.0153	.0096
%RSD	71.85	42.75	58.03

#1	-.0083	-.0220	-.0038
#2	-.0056	-.0295	-.0131
#3	-.0080	-.0487	-.0222
#4	-.0134	-.0234	-.0144
#5	-.0283	-.0555	-.0291

Int. Std.	In ₁₁₅
Line	115/pulse
Units	Cts/S
Avg	11.368
Stddev	.087
%RSD	.76526

#1	11.410
#2	11.383
#3	11.236
#4	11.342
#5	11.469

Method: I-129

Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/22/99

10:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1405	-.0288	-.0176
Stddev	.0085	.0244	.0272
%RSD	6.064	84.75	154.4
#1	-.1397	-.0646	-.0510
#2	-.1331	-.0376	-.0407
#3	-.1475	-.0060	.0035
#4	-.1316	-.0291	-.0108
#5	-.1508	-.0065	.0110

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.311
Stddev	.845
%RSD	7.4728

#1	12.726
#2	11.454
#3	10.680
#4	10.923
#5	10.775

Method: I-129   Sample Name: s98t003389   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 10:37 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe [131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0559	.0035	.0050
Stddev	.0133	.0160	.0167
%RSD	23.73	453.3	335.8

#1	.0656	.0248	.0075
#2	.0732	-.0129	.0326
#3	.0533	.0084	-.0083
#4	.0440	-.0121	-.0072
#5	.0433	.0096	.0003

Int. Std.	In[115]
Line	115/pulse
Units	Cts/s
Avg	11.020
Stddev	.107
%RSD	.97297

#1	10.840
#2	11.070
#3	11.051
#4	11.020
#5	11.120

3359

Method: I-129   Sample Name: s98t003389_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 10:38 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0399	-.0037	.0080
Stddev	.0080	.0206	.0179
%RSD	19.94	555.7	223.2

#1	.0353	-.0048	.0150
#2	.0378	.0103	.0050
#3	.0514	.0141	.0297
#4	.0310	-.0380	-.0193
#5	.0440	-.0001	.0097

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.148
Stddev	.083
%RSD	.74678

#1	11.151
#2	11.150
#3	11.022
#4	11.256
#5	11.162

Method: I-129

Sample Name: s98t003394

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 10:40 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0299	.0004	.0135
Stddev	.0084	.0096	.0153
%RSD	28.14	2698.	113.2
#1	.0244	.0051	.0238
#2	.0185	-.0111	-.0063
#3	.0350	.0122	.0155
#4	.0320	-.0077	.0031
#5	.0396	.0034	.0315

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.378
Stddev	.096
%RSD	.84098
#1	11.495
#2	11.430
#3	11.270
#4	11.289
#5	11.407

Method: I-129   Sample Name: s98t003394_d   Operator: bjg
Comment: 201 df
Run Time: 03/22/99 10:41 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0302	.0005	.0061
Stddev	.0117	.0106	.0088
%RSD	38.67	2271.	143.6

#1	.0238	.0063	.0157
#2	.0316	.0035	.0106
#3	.0335	.0025	.0093
#4	.0154	-.0180	-.0068
#5	.0468	.0081	.0018

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.383
Stddev	.098
%RSD	.86521

#1	11.429
#2	11.495
#3	11.356
#4	11.403
#5	11.231

HNF-1663 REV. 0

Method: I-129   Sample Name: s98t003399   Operator: bjg
Comment: 20301 df
Run Time: 03/22/99 10:43 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0275	-.0143	-.0024
Stddev	.0063	.0223	.0102
%RSD	23.11	155.8	432.3

#1	.0245	-.0081	.0098
#2	.0250	-.0536	-.0135
#3	.0384	-.0011	-.0109
#4	.0223	-.0008	.0062
#5	.0271	-.0078	-.0034

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.643
Stddev	.126
%RSD	1.0855

#1	11.682
#2	11.821
#3	11.647
#4	11.475
#5	11.593

3363

Method: I-129

Sample Name: s98t003399_d

Operator: bjg

Comment: 20301 df

Run Time: 03/22/99

10:44 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0216	-.0186	-.0091
Stddev	.0111	.0068	.0101
%RSD	51.60	36.67	111.6

#1	.0273	-.0181	-.0212
#2	.0355	-.0199	.0054
#3	.0111	-.0175	-.0041
#4	.0246	-.0091	-.0134
#5	.0093	-.0282	-.0121

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.690
Stddev	.028
%RSD	.24043

#1	11.699
#2	11.655
#3	11.667
#4	11.710
#5	11.720

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 10:47 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.034	10.10	.0552
Stddev	.044	.24	.0060
%RSD	2.164	2.336	10.84
#1	2.060	10.03	.0485
#2	2.055	10.23	.0510
#3	2.076	10.44	.0596
#4	1.969	9.970	.0539
#5	2.009	9.831	.0629

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.9526
Stddev	.1730
%RSD	1.9329

#1	9.2180
#2	8.9202
#3	8.7538
#4	8.8716
#5	8.9993

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 10:49 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0043	-.0282	-.0139
Stddev	.0097	.0146	.0148
%RSD	224.2	51.84	106.8

#1	-.0058	-.0095	.0081
#2	.0104	-.0313	-.0306
#3	-.0015	-.0253	-.0072
#4	-.0154	-.0250	-.0188
#5	-.0093	-.0500	-.0209

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.059
Stddev	.114
%RSD	.94281

#1	11.906
#2	11.974
#3	12.135
#4	12.103
#5	12.174

LABCORE Completed Worklist Report for Worklist# 28845

Analyst: bjg

Instrument: ICPMS1

Book#: 990319-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) I-129

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV		0	I129	QC	0.00927	9.88e-03	106.580	% Recovery
2	ICB		0	I129	QC	1	<1.20e-4		ug/mL
3	SAMPLE	S98T002809	0 D	I129-D1	LIQUID	<u>N/A</u>	4.860e+00	4.860	ug/mL
4	DUP	S98T002809	0 D	I129-D1	LIQUID	<4.86e0	<4.86e0		RPD
5	SAMPLE	S98T002968	0 D	I129-D1	LIQUID	<u>N/A</u>	4.860e+00	4.860	ug/mL
6	DUP	S98T002968	0 D	I129-D1	LIQUID	<4.86e0	<4.86e0		RPD
7	CCV		0	I129	QC	0.00927	9.83e-03	106.041	% Recovery
8	CCB		0	I129	QC	1	<1.20e-4		ug/mL

Final page for worklist# 28845

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Sail M. Pang 03/22/99
 Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28845

Analyst: BJL Instrument: ICPMS1 Book# 890319-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (DIRECT) I-129

S Type	Sample#	R A	Test	Matrix	Group#	Project
1	ICV		II129	QC		
2	ICB		II129	QC		
3	SAMPLE	S98T002809 0 D	II129-D1	LIQUID	98000437	BY-105 (2)
Analytes Requested: II129-D1						
4	DUP	S98T002809 0 D	II129-D1	LIQUID		
5	SAMPLE	S98T002968 0 D	II129-D1	LIQUID	98000509	BY-105 (2)
Analytes Requested: II129-D1						
6	DUP	S98T002968 0 D	II129-D1	LIQUID		
7	~~SAMPLE~~	~~S98T002968 0 D~~	~~II129-D1~~	~~LIQUID~~		
8	CCV		II129	QC		
9	CCB		II129	QC		

Final page for worklist # 28845

Blaxch 3-19-99
Signature Date

S98T002809
2809-D
2968
2968-D

1-10-1 - 1021-10-1

Signature Date

DF
40,507

Data Entry Comments:

Uploaded + Validated by:

Sail M. Pay

03/22/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3368

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/19/99

11:20 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	119900.
Stddev	485.
%RSD	.4046

#1	119300.
#2	120000.
#3	120100.
#4	120300.
#5	119500.
#6	119000.
#7	119900.
#8	120400.
#9	119700.
#10	120400.

3-19-99

BY-105 F-125

598TC02809

2968

2604

2605

2606

File # 10315A.TVF

W.C. # 28845

28546

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3364 TO 3381.

Method: I-129 Sample Name: Blank

Operator:

Comment:

Run Time: 03/19/99 11:22 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	69.81	77.25	63.80
Stddev	2.04	1.68	2.63
%RSD	2.927	2.178	4.116

#1	71.01	78.72	66.98
#2	70.49	75.97	65.54
#3	67.24	78.41	64.09
#4	68.15	78.21	61.83
#5	72.14	74.96	60.57

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.7805
Stddev	.1157
%RSD	1.3171

#1	8.6890
#2	8.6815
#3	8.7299
#4	8.9440
#5	8.8581

Method : I-129

Standardzn Report

03/19/99 11:24:08

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	204.2119	69.8058	1.0000000	03/19/99 11:23:22
I_[129]	203.7662	77.2529	1.0000000	03/19/99 11:23:22
Xe[131]	204.5712	63.8018	1.0000000	03/19/99 11:23:22

3371

Analysis Report

03/19/99 11:26:32

page 1

Method: I-129

Sample Name: I-127

Operator:

Comment:

Run Time: 03/19/99 11:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I ₋ [127]	I ₋ [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3550.	3550.	3550.
Stddev	45.	45.	45.
%RSD	1.257	1.257	1.257

#1	3493.	3493.	3493.
#2	3565.	3565.	3565.
#3	3602.	3602.	3602.
#4	3514.	3514.	3514.
#5	3574.	3574.	3574.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.8265
Stddev	.0536
%RSD	.60748

#1	8.8228
#2	8.8951
#3	8.7451
#4	8.8369
#5	8.8328

3372

Method : I-129

Standardzn Report

03/19/99 11:26:39

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.2443	69.8058	1.0000000	03/19/99 11:26:30
I_[129]	207.7987	77.2529	1.0000000	03/19/99 11:26:30
Xe[131]	208.6036	63.8018	1.0000000	03/19/99 11:26:30

3373

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/19/99 11:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.058	9.883	.0527
Stddev	.103	.501	.0301
%RSD	4.987	5.067	57.09

#1	1.990	9.476	.0301
#2	1.968	9.525	.0243
#3	2.037	9.947	.0555
#4	2.228	10.71	.1006
#5	2.069	9.753	.0529

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	6.7916
Stddev	.5068
%RSD	7.4625

#1	7.3704
#2	7.1077
#3	6.8710
#4	6.0684
#5	6.5403

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmah

Run Time: 03/19/99 11:35 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0128	.0022	-.0014
Stddev	.0033	.0076	.0058
%RSD	25.69	340.1	426.6

#1	-.0138	.0116	.0087
#2	-.0153	-.0026	-.0054
#3	-.0143	.0092	-.0022
#4	-.0070	-.0050	-.0052
#5	-.0135	-.0021	-.0027

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.0815
Stddev	.1865
%RSD	2.0533

#1	9.0385
#2	9.1189
#3	8.7845
#4	9.2626
#5	9.2032

Method: I-129

Sample Name: s98t002809

Operator: bjg

Comment: 40501 df

Run Time: 03/19/99

11:38 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0033	.0205	-.0062
Stddev	.0110	.0117	.0078
%RSD	329.5	57.19	127.3
#1	.0093	.0264	-.0033
#2	.0018	.0363	-.0051
#3	-.0059	.0071	-.0028
#4	-.0204	.0216	-.0198
#5	-.0016	.0112	.0002

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.1837
Stddev	.1249
%RSD	1.3597

#1	8.9914
#2	9.1397
#3	9.3096
#4	9.2663
#5	9.2115

Method: I-129   Sample Name: s98t002809_d   Operator: bjg
Comment: 40501 df
Run Time: 03/19/99 11:39 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0018	.0122	-.0130
Stddev	.0107	.0127	.0091
%RSD	607.8	104.0	70.34

#1	-.0128	.0087	-.0185
#2	.0022	.0040	-.0156
#3	-.0030	.0015	-.0218
#4	.0062	.0137	-.0104
#5	.0162	.0334	.0015

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.3691
Stddev	.1227
%RSD	1.3096

#1	9.4491
#2	9.4714
#3	9.3673
#4	9.3950
#5	9.1625

Method: I-129 Sample Name: s98t002968

Operator: bjg

Comment: 40501 df

Run Time: 03/19/99 11:41 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0048	.0188	-.0098
Stddev	.0140	.0076	.0102
%RSD	294.0	40.72	103.6

#1	-.0026	.0086	-.0178
#2	-.0221	.0206	-.0228
#3	.0031	.0171	-.0080
#4	.0130	.0177	.0001
#5	-.0153	.0298	-.0007

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.4219
Stddev	.0683
%RSD	.72489

#1	9.4703
#2	9.4996
#3	9.3979
#4	9.3235
#5	9.4184

Method: I-129   Sample Name: s98t002968_d   Operator: bjg
Comment: 40501 df
Run Time: 03/19/99 11:43 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0024	.0134	-.0068
Stddev	.0089	.0131	.0108
%RSD	370.3	98.20	158.8

#1	-.0119	.0150	-.0014
#2	-.0103	.0321	-.0106
#3	.0097	-.0030	-.0183
#4	.0016	.0060	.0091
#5	-.0012	.0168	-.0128

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.5067
Stddev	.1450
%RSD	1.5255

#1	9.3137
#2	9.5633
#3	9.7100
#4	9.4662
#5	9.4802

Method: I-129 Sample Name: c6v

Operator: bjg

Comment: itank

Run Time: 03/19/99 12:11 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I [127]	I [129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.009	9.830	.0355
Stddev	.064	.297	.0138
%RSD	3.187	3.018	38.86
#1	1.914	9.387	.0385
#2	2.004	9.749	.0158
#3	2.063	9.870	.0456
#4	1.990	9.953	.0277
#5	2.073	10.19	.0497

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	7.4248
Stddev	.2940
%RSD	3.9601

#1	7.9065
#2	7.5033
#3	7.2569
#4	7.2646
#5	7.1929

Method: I-129

Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/19/99 12:15

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0023	.0013	-.0009
Stddev	.0074	.0093	.0119
%RSD	326.6	704.7	1262.

#1	.0040	.0066	.0089
#2	-.0145	-.0136	-.0166
#3	-.0035	-.0013	-.0049
#4	.0023	.0050	-.0052
#5	.0004	.0099	.0130

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.8998
Stddev	.0839
%RSD	.84799

#1	9.8380
#2	10.010
#3	9.9301
#4	9.7965
#5	9.9240

LABCORE Completed Worklist Report for Worklist# 28912

Analyst: bjg

Instrument: ICPMS1

Book#: 990319-3

Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (Solid Acid Digest)

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1	ICV	0	I129	QC	0.00927	1.01e-02	108.954 % Recovery	
2	ICB	0	I129	QC	1	<1.20e-4	ug/mL	
3	BLNK-PREP	0	I129-A1	SOLID	1	<1.20e-3	ug/g	
4	SAMPLE	S98T002607 0 A	I129-A1	SOLID	N/A <	4.683e+00	4.683	ug/g
5	DUP	S98T002607 0 A	I129-A1	SOLID	<4.68e0	<4.82e0	RPD	
6	SAMPLE	S98T002608 0 A	I129-A1	SOLID	N/A <	4.818e+02	481.700	ug/g
7	DUP	S98T002608 0 A	I129-A1	SOLID	<4.82e2	<4.71e2	RPD	
8	SAMPLE	S98T002609 0 A	I129-A1	SOLID	N/A <	4.955e+02	495.500	ug/g
9	DUP	S98T002609 0 A	I129-A1	SOLID	<4.96e2	<4.66e2	RPD	
10	CCV	0	I129	QC	0.00927	9.93e-03	107.120 % Recovery	
11	CCB	0	I129	QC	1	<1.20e-4	ug/mL	

Final page for worklist# 28912

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Saul M. Pang 03/23/99
Reviewer Signature Date

LABCORE Data Entry Template for Worklist# 28912

Analyst: BJG Instrument: ICPMS1 Book# 990319-3Method: LA-506-101 Rev/Mod A-0

Worklist Comment: ICP/MS BY-105 (Solid Acid Digest)

S Type	Sample#	R	A	Test	Matrix	Group#	Project
1 ICV				I129	QC		
2 ICB				I129	QC		
3 BLNK-PREP				I129-A1	SOLID		
4 SAMPLE	S98T002607	0	A	I129-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: I129-A1							
5 DUP	S98T002607	0	A	I129-A1	SOLID		
6 SAMPLE	S98T002608	0	A	I129-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: I129-A1							
7 DUP	S98T002608	0	A	I129-A1	SOLID		
8 SAMPLE	S98T002609	0	A	I129-A1	SOLID	98000436	BY-105 (2)
Analytes Requested: I129-A1							
9 DUP	S98T002609	0	A	I129-A1	SOLID		
10 CCV				I129	QC		
11 CCB				I129	QC		

Final page for worklist # 28912

	<u>3-22-99</u>	<u>Signature</u>	<u>Date</u>
PREP BLK	1ml - 9ml	DF.	
S98T002607	0.05ml - 10ml	10	
2607-1	↓	201	
2608	0.05ml - 10ml - 1ml - 10ml	201	
2608-1	↓	20,301	
2608			
2608-1			
Data Entry Comments			

Uploaded + Validated by:

 03/23/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3383

Analysis Report

HNF-1663 REV. 0

03/22/99 10:22:19

page 1

Method: MS_TUNE

Sample Name: ms-tune

Operator: bjg

Comment:

Run Time: 03/22/99 10:21 Type: Unk

Mode: INT

Corr.Fact: 1.000000

Elem	In[115]
Line	115/pulse
Units	Cts/S
Avg	140300.
Stddev	1558.
%RSD	1.111

#1	142200.
#2	141000.
#3	139800.
#4	139900.
#5	136400.
#6	140000.
#7	141200.
#8	141500.
#9	140300.
#10	140300.

3-22-99
BY-105
598T02607-09
FILE # 10322A.TXT
W.L.H 28812

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3384 TO 3400

3384

Method: I-129 Sample Name: Blank

Operator:

Comment:

Run Time: 03/22/99 10:23 Type: Std Mode: IR Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	70.71	144.5	118.3
Stddev	.97	3.4	4.0
%RSD	1.378	2.361	3.408

#1	71.77	149.3	124.0
#2	70.70	145.8	117.6
#3	69.58	140.0	114.1
#4	71.57	143.5	115.2
#5	69.91	143.8	120.5

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.703
Stddev	.136
%RSD	1.2739

#1	10.533
#2	10.587
#3	10.843
#4	10.807
#5	10.743

Method : I-129

Standardzn Report

03/22/99 10:24:47

page 1

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	208.1904	70.7074	1.0000000	03/22/99 10:24:30
I_[129]	203.7753	144.4827	1.0000000	03/22/99 10:24:30
Xe[131]	205.3432	118.2832	1.0000000	03/22/99 10:24:30

3386

Analysis Report

HNF-1663 REV 0
03/22/99 10:26:44

page 1

Method: I-129 Sample Name: I-127

Operator:

Comment:

Run Time: 03/22/99 10:25 Type: Std

Mode: IR

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	Cts/S	Cts/S	Cts/S
Avg	3806.	3806.	3806.
Stddev	50.	50.	50.
%RSD	1.317	1.317	1.317
#1	3727.	3727.	3727.
#2	3791.	3791.	3791.
#3	3819.	3819.	3819.
#4	3848.	3848.	3848.
#5	3847.	3847.	3847.

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	10.912
Stddev	.096
%RSD	.88085

#1	11.030
#2	10.944
#3	10.922
#4	10.899
#5	10.764

3387

El Name	Slope	Y-int	Correlation	Date Stdized
I_[127]	223.5636	70.7074	1.0000000	03/22/99 10:26:42
I_[129]	219.1486	144.4827	1.0000000	03/22/99 10:26:42
Xe[131]	220.7165	118.2832	1.0000000	03/22/99 10:26:42

3388

Method: I-129 Sample Name: blank

Operator: bjg

Comment: background

Run Time: 03/22/99 10:29 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0029	-.0012	-.0058
Stddev	.0080	.0076	.0132
%RSD	274.6	635.4	227.4
#1	.0096	.0017	.0103
#2	-.0052	-.0037	-.0108
#3	-.0056	.0051	-.0081
#4	.0115	.0042	.0033
#5	.0043	-.0133	-.0237

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.109
Stddev	.080
%RSD	.71806

#1	11.181
#2	11.178
#3	11.129
#4	11.062
#5	10.995

Method: I-129 Sample Name: icv

Operator: bjg

Comment: itank

Run Time: 03/22/99 10:31 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	1.999	10.06	.0465
Stddev	.049	.33	.0147
%RSD	2.448	3.291	31.71
#1	2.053	10.54	.0661
#2	1.934	9.930	.0465
#3	1.962	9.633	.0330
#4	2.029	10.03	.0552
#5	2.015	10.14	.0315

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	8.3301
Stddev	.1762
%RSD	2.1149
#1	8.0639
#2	8.4749
#3	8.5056
#4	8.3092
#5	8.2969

Method: I-129 Sample Name: icb

Operator: bjg

Comment: 2% tmaH

Run Time: 03/22/99 10:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0127	-.0358	-.0165
Stddev	.0091	.0153	.0096
%RSD	71.85	42.75	58.03
#1	-.0083	-.0220	-.0038
#2	-.0056	-.0295	-.0131
#3	-.0080	-.0487	-.0222
#4	-.0134	-.0234	-.0144
#5	-.0283	-.0555	-.0291

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	11.368
Stddev	.087
%RSD	.76526

#1	11.410
#2	11.383
#3	11.236
#4	11.342
#5	11.469

Method: I-129 Sample Name: prepblk

Operator: bjg

Comment: 10 df

Run Time: 03/22/99 11:24 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.1589	-.0109	-.0183
Stddev	.0044	.0137	.0152
%RSD	2.790	125.9	82.79

#1	-.1593	-.0152	-.0251
#2	-.1586	-.0210	-.0387
#3	-.1644	.0078	-.0169
#4	-.1521	-.0013	.0025
#5	-.1603	-.0247	-.0134

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.024
Stddev	.234
%RSD	1.9489

#1	12.164
#2	12.242
#3	11.797
#4	11.744
#5	12.174

Method: I-129

Sample Name: s98t002607

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 11:26

Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_ ₁₂₇	I_ ₁₂₉	Xe ₁₃₁
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0148	.0056	-.0002
Stddev	.0160	.0176	.0182
%RSD	108.4	311.1	7675.

#1	.0340	.0120	.0187
#2	.0204	-.0002	-.0176
#3	-.0103	-.0170	-.0051
#4	.0157	.0026	.0191
#5	.0143	.0309	-.0164

Int. Std.	In ₁₁₅
Line	115/pulse
Units	Cts/S
Avg	12.032
Stddev	.086
%RSD	.71610

#1	11.984
#2	12.019
#3	11.957
#4	12.020
#5	12.179

Method: I-129

Sample Name: s98t002607_d

Operator: bjg

Comment: 201 df

Run Time: 03/22/99 11:27 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0027	.0124	.0046
Stddev	.0012	.0120	.0141
%RSD	44.79	96.66	305.6

#1	-.0021	.0160	-.0051
#2	-.0028	.0108	.0054
#3	-.0011	.0179	.0165
#4	-.0028	.0244	.0199
#5	-.0044	-.0072	-.0136

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.159
Stddev	.068
%RSD	.56185

#1	12.103
#2	12.187
#3	12.200
#4	12.073
#5	12.235

Method: I-129   Sample Name: s98t002608   Operator: bjg
Comment: 20301 df
Run Time: 03/22/99 11:29 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0245	-.0193	-.0190
Stddev	.0079	.0124	.0110
%RSD	32.23	64.47	57.90

#1	-.0371	-.0175	-.0243
#2	-.0210	-.0135	-.0069
#3	-.0171	-.0220	-.0123
#4	-.0269	-.0050	-.0164
#5	-.0203	-.0385	-.0351

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.734
Stddev	.088
%RSD	.68842

#1	12.851
#2	12.655
#3	12.785
#4	12.738
#5	12.643

Method: I-129

Sample Name: s98t002608_d

Operator: bjg

Comment: 20301 df

Run Time: 03/22/99 11:30 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0232	-.0139	-.0083
Stddev	.0039	.0122	.0131
%RSD	16.81	87.89	158.3

#1	-.0186	-.0158	-.0094
#2	-.0282	.0037	-.0147
#3	-.0253	-.0234	-.0134
#4	-.0239	-.0077	.0145
#5	-.0200	-.0265	-.0185

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.690
Stddev	.073
%RSD	.57343

#1	12.736
#2	12.678
#3	12.611
#4	12.636
#5	12.789

Method: I-129   Sample Name: s98t002609   Operator: bjg
Comment: 20301 df
Run Time: 03/22/99 11:32 Type: Unk   Mode: CONC   Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0157	-.0185	-.0198
Stddev	.0071	.0130	.0106
%RSD	45.19	70.38	53.22

#1	-.0151	-.0167	-.0061
#2	-.0165	-.0184	-.0348
#3	-.0082	.0019	-.0212
#4	-.0117	-.0271	-.0148
#5	-.0270	-.0321	-.0223

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.771
Stddev	.090
%RSD	.70085

#1	12.801
#2	12.752
#3	12.702
#4	12.690
#5	12.911

Method: I-129

Sample Name: s98t002609_d

Operator: bjg

Comment: 20301 df

Run Time: 03/22/99 11:33 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	-.0209	-.0298	-.0281
Stddev	.0066	.0106	.0080
%RSD	31.82	35.77	28.31

#1	-.0122	-.0414	-.0321
#2	-.0284	-.0175	-.0150
#3	-.0265	-.0271	-.0299
#4	-.0194	-.0402	-.0275
#5	-.0177	-.0226	-.0360

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.755
Stddev	.089
%RSD	.69616

#1	12.901
#2	12.712
#3	12.665
#4	12.740
#5	12.757

Method: I-129 Sample Name: ccv

Operator: bjg

Comment: itank

Run Time: 03/22/99 11:35 Type: Unk

Mode: CONC

Corr.Fact.: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	2.004	9.932	.0427
Stddev	.074	.236	.0110
%RSD	3.694	2.377	25.70
#1	1.881	9.579	.0515
#2	2.068	10.15	.0528
#3	1.996	9.823	.0256
#4	2.031	9.983	.0400
#5	2.047	10.12	.0437

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	9.6881
Stddev	.3481
%RSD	3.5934
#1	10.223
#2	9.6624
#3	9.7720
#4	9.4766
#5	9.3063

Method: I-129 Sample Name: ccb

Operator: bjg

Comment: 2% tmah

Run Time: 03/22/99 11:37 Type: Unk

Mode: CONC

Corr.Fact: 1.000000

Elem	I_[127]	I_[129]	Xe[131]
Line	127/pulse	129/pulse	131/pulse
Units	ppb	ppb	ppb
Avg	.0054	-.0281	-.0242
Stddev	.0064	.0058	.0072
%RSD	118.6	20.71	30.00

#1	.0118	-.0278	-.0159
#2	-.0054	-.0233	-.0343
#3	.0077	-.0231	-.0279
#4	.0058	-.0290	-.0234
#5	.0072	-.0374	-.0192

Int. Std.	In[115]
Line	115/pulse
Units	Cts/S
Avg	12.706
Stddev	.116
%RSD	.91482

#1	12.508
#2	12.696
#3	12.763
#4	12.778
#5	12.785

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 11 of 17

Document Information

Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 3401 THRU 3673)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	TOTAL INORGANIC CARBON/TOTAL ORGANIC CARBON ANALYSIS, TICTOC WORKLIST 26744 THRU 26753		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

HNF-1663 REV. 0

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

TABLE OF CONTENTS (Continued)

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 5
of 17
Cont.Sec 6
of 17Sec 7
of 17

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes	5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26744

Analyst: rdm

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTs

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@TICTOC1 TIC-02	LIQUID	6.02E+02	5.42E+2	90.033 % Recovery	
1 STD	0	@TICTOC1 TOC-02	LIQUID	3.00E+03	2.80E+3	93.333 % Recovery	
2 BLNK	0	@TICTOC1 TIC-02	LIQUID	1	3.40E+0	3.400 ug/mL	
2 BLNK	0	@TICTOC1 TOC-02	LIQUID	1	3.20E+0	3.200 ug/mL	
3 SAMPLE	S98T002392 0	@TICTOC1 TIC-02	LIQUID	N/A	3.15E+03	5.000 ug/mL	
3 SAMPLE	S98T002392 0	@TICTOC1 TOC-02	LIQUID	N/A	4.00E+01	40.000 ug/mL	
4 DUP	S98T002392 0	@TICTOC1 TIC-02	LIQUID	3.15E+1	3.11E+1	1.278 RPD	
4 DUP	S98T002392 0	@TICTOC1 TOC-02	LIQUID	<4.00E+1	<4.00E+1	RPD	
5 SAMPLE	S98T002822 0	@TICTOC1 TIC-02	LIQUID	N/A	7.68E+03	5.000 ug/mL	
5 SAMPLE	S98T002822 0	@TICTOC1 TOC-02	LIQUID	N/A	2.13E+3	40.000 ug/mL	
6 DUP	S98T002822 0	@TICTOC1 TIC-02	LIQUID	7.68E+3	7.75E+3	0.907 RPD	
6 DUP	S98T002822 0	@TICTOC1 TOC-02	LIQUID	2.13E+3	2.20E+3	3.233 RPD	
7 SAMPLE	S98T002823 0	@TICTOC1 TIC-02	LIQUID	N/A	5.67E+03	5.000 ug/mL	
7 SAMPLE	S98T002823 0	@TICTOC1 TOC-02	LIQUID	N/A	2.31E+3	40.000 ug/mL	
8 DUP	S98T002823 0	@TICTOC1 TIC-02	LIQUID	5.67E+3	5.53E+3	2.500 RPD	
8 DUP	S98T002823 0	@TICTOC1 TOC-02	LIQUID	2.31E+3	2.03E+3	12.903 RPD	

Final page for worklist# 26744

Analyst Signature _____ Date _____

Ch. Pinner 12/1/98
Analyst Signature _____ Date _____

[Signature] 12/2/98
Reviewer Signature _____ Date _____

10/23/98 09:33

40004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26744

Analyst: Rm Instrument: CARB2 Book# 25N120
26N12F
 Method: LA-342-100 Rev/Mod E-1
 Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	LIQUID		
2 BLNK			@TICTOC1	LIQUID		
3 SAMPLE	S98T002392 0		@TICTOC1	LIQUID	98000436	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02 <i>Steer 3/23/99</i>						
4 DUP	S98T002392 0		@TICTOC1	LIQUID		
5 SAMPLE	S98T002822 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02						
6 DUP	S98T002822 0		@TICTOC1	LIQUID		
7 SAMPLE	S98T002823 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02						
8 DUP	S98T002823 0		@TICTOC1	LIQUID		

Final page for worklist # 26744

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3402

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	555.1	593.1
26744	μg of Carbon from Baseline (C2)	13.2	33.8
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98005345			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
12:00 AM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.42E+02	2.80E+03
BY-105	Percent Standard Recovery	90.0	93.2

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>WA</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3403

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	9.8	30.6
26744	µg of Carbon from Baseline (C2)	13.2	33.8

Test Code
@TIC/TOC1
Matrix
LIQUID
Batch Number
98005345
Rerun
0
Sample Prep
N/A
Sample #
BLANK
Instrument Code
CARB2
Prepared By
CJO
Chemist
MJL
Analyst
RDM
Date Complete
12/01/98
Analysis Date
11/28/98
Analysis Time
12:00 AM
Sample Point
BY-105

µg of Carbon = |C1-C2|

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
12:00 AM			
Sample Point	µg of Carbon	3.40E+00	3.20E+00
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3404

WORKBOOK PAGE: SAM5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	781	246.6
26744	µg of Carbon from Baseline (C2)	13.2	33.8
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005345			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002822			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
12:00 AM		5	40
Sample Point	µg of Carbon/mL	7.68E+03	2.13E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3405

WORKBOOK PAGE: DUP6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	788.6	254.2
26744	µg of Carbon from Baseline (C2)	13.2	33.8
Test Code	Known µg of C from Original Sample	7.68E+3	2.13E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005345			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002822			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
12:00 AM			
Sample Point	µg of Carbon/mL	7.75E+03	2.20E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SAM7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	579.8	264.5
26744	µg of Carbon from Baseline (C2)	13.2	33.8
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005345			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002823			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
12:00 AM		5	40
Sample Point	µg of Carbon/mL	5.67E+03	2.31E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3407

WORKBOOK PAGE: DUP8

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	565.8	237.2
26744	µg of Carbon from Baseline (C2)	13.2	33.8
Test Code	Known µg of C from Original Sample	5.67E+3	2.31E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005345			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002823			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
12:00 AM			
Sample Point	µg of Carbon/mL	5.53E+03	2.03E+03
BY-105			

Data Entered By:

CJO

Date:

12/01/98

Signature of Chemist:

NA

Date:

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/27/98

Time: 19:59:06

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE

Blank Value = N/A

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	1.50	0.00
2	1.00	2.70	44.44
3	1.50	4.10	34.15
4	2.00	5.00	18.00
5	2.50	5.80	13.79
6	3.00	6.50	10.77
7	3.50	7.00	7.14
8	4.00	7.50	6.67
9	4.50	8.00	6.25
10	5.00	8.40	4.76
11	5.50	8.90	5.62
12	6.00	9.30	4.30
13	6.50	9.70	4.12
14	7.00	10.10	3.96
15	7.50	10.50	3.81
16	8.00	10.90	3.67
17	8.50	11.30	3.54
18	9.00	11.70	3.42
19	9.50	12.10	3.31
20	10.00	12.40	2.42
21	10.50	12.80	3.13
22	11.00	13.20	3.03

BLANK VALUE = 13.2 micrograms carbon

BLANK FACTOR = 13.2 / 10.9989 =

+1.20E+00

ug/min Carbon

Sample Run By:

RD MEYERS

00002

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ ON PAGES 3429 TO 3426

3409

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TIC/TOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/27/98

Time: 20:12:32

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE

Blank Value = N/A

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	2.70	88.89
3	1.50	9.40	71.28
4	2.00	15.70	40.13
5	2.50	20.50	23.41
6	3.00	23.60	13.14
7	3.50	25.30	6.72
8	4.00	26.40	4.17
9	4.50	27.40	3.65
10	5.00	28.10	2.49
11	5.50	28.80	2.43
12	6.00	29.40	2.04
13	6.50	29.90	1.67
14	7.00	30.40	1.64
15	7.50	30.90	1.62
16	8.00	31.40	1.59
17	8.50	31.90	1.57
18	9.00	32.40	1.54
19	9.50	32.80	1.22
20	10.00	33.10	0.91
21	10.50	33.50	1.19
22	11.00	33.80	0.89

BLANK VALUE = 33.8 micrograms carbon

BLANK FACTOR = 33.8 / 10.99597 =

+3.07E+00

ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3410

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

Sample: 25N12-D

Date: 11/27/98

Time: 20:27:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	34.90	99.14
3	1.51	170.30	79.51
4	2.01	304.80	44.13
5	2.50	410.20	25.69
6	3.00	478.60	14.29
7	3.50	517.40	7.50
8	4.00	536.20	3.51
9	4.50	544.50	1.52
10	5.00	547.90	0.62
11	5.50	549.70	0.33
12	6.00	550.70	0.18
13	6.50	551.30	0.11
14	7.00	551.90	0.11
15	7.50	552.30	0.07
16	8.00	552.80	0.09
17	8.50	553.30	0.09
18	9.00	553.70	0.07
19	9.50	554.00	0.05
20	10.00	554.40	0.07
21	10.50	554.70	0.05
22	11.00	555.10	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 =

+1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(555.1 - 13.19854) (1)/(1) =

+5.419E+02 g/L Carbon

(555.1 - 13.19854) (1)/(1) (12) =

+4.516E+01 Molar Carbon

Sample Run By:

RD MEYERS

00002

3411

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 26N12-F

Date: 11/27/98

Time: 20:40:37

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.30	0.00
2	1.00	13.40	97.76
3	1.50	80.00	83.25
4	2.00	234.80	65.93
5	2.50	413.50	43.22
6	3.00	515.70	19.82
7	3.50	558.00	7.58
8	4.00	574.70	2.91
9	4.50	581.20	1.12
10	5.00	584.30	0.53
11	5.50	586.10	0.31
12	6.00	587.20	0.19
13	6.50	588.20	0.17
14	7.00	589.00	0.14
15	7.50	589.60	0.10
16	8.00	590.30	0.12
17	8.50	590.80	0.08
18	9.00	591.30	0.08
19	9.50	591.70	0.07
20	10.00	592.20	0.08
21	10.50	592.70	0.08
22	11.00	593.10	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(593.1 - 33.76101) (1)/(1) = +5.593E+02 g/L Carbon
 (593.1 - 33.76101) (1)/(1) (12) = +4.661E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3412

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2-0

Sample: BLK

Date: 11/27/98

Time: 20:53:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.30	69.23
3	1.50	2.80	53.57
4	2.00	3.70	24.32
5	2.50	4.40	15.91
6	3.00	4.80	8.33
7	3.50	5.20	7.69
8	4.00	5.60	7.14
9	4.50	5.80	3.45
10	5.00	6.20	6.45
11	5.50	6.50	4.62
12	6.00	6.80	4.41
13	6.50	7.10	4.23
14	7.00	7.40	4.05
15	7.50	7.70	3.90
16	8.00	8.00	3.75
17	8.50	8.30	3.61
18	9.00	8.60	3.49
19	9.50	8.90	3.37
20	10.00	9.20	3.26
21	10.50	9.50	3.16
22	11.00	9.80	3.06

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 = +1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(9.8 - 13.19751) (1)/(1) = < 5.00 E-3 g/L Carbon
 (9.8 - 13.19751) (1)/(1) (12) = < 4.17 E-4 Molar Carbon

Sample Run By:

RD MEYERS

00002

3413

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/27/98

Time: 22:11:33

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.40	0.00
2	1.01	4.70	91.49
3	1.50	11.40	58.77
4	2.00	16.50	30.91
5	2.50	20.10	17.91
6	3.00	22.30	9.87
7	3.50	23.50	5.11
8	4.00	24.50	4.08
9	4.50	25.30	3.16
10	5.00	26.00	2.69
11	5.50	26.40	1.52
12	6.00	26.90	1.86
13	6.50	27.40	1.82
14	7.00	27.70	1.08
15	7.50	28.10	1.42
16	8.00	28.50	1.40
17	8.50	28.90	1.38
18	9.00	29.20	1.03
19	9.50	29.60	1.35
20	10.00	29.90	1.00
21	10.50	30.30	1.32
22	11.00	30.60	0.98

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(30.6 - 33.76625) (1)/(1) = < 5.00 E-3 g/L Carbon
 (30.6 - 33.76625) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3414

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2392

Date: 11/27/98

Time: 22:39:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.20	0.00
2	1.01	1.10	81.82
3	1.50	6.50	83.08
4	2.00	17.30	62.43
5	2.50	26.80	35.45
6	3.00	33.40	19.76
7	3.50	37.70	11.41
8	4.00	40.00	5.75
9	4.50	41.00	2.44
10	5.00	41.70	1.68
11	5.50	42.00	0.71
12	6.00	42.30	0.71
13	6.50	42.60	0.70
14	7.00	42.80	0.47
15	7.50	43.10	0.70
16	8.00	43.30	0.46
17	8.50	43.60	0.69
18	9.00	43.80	0.46
19	9.50	44.00	0.45
20	10.00	44.20	0.45
21	10.50	44.40	0.45
22	11.00	44.70	0.67

**DATA
NOT USED**

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 = +1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(44.7 - 13.19751) (1)/(1) = +3.15E+01 g/L Carbon
 (44.7 - 13.19751) (1)/(1) (12) = +2.63E+00 Molar Carbon

Sample Run By:

RD MEYERS

00002

3415

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 210

Sample: 2392

Date: 11/27/98

Time: 22:52:07

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	1.10	72.73
3	1.51	6.30	82.54
4	2.01	10.80	41.67
5	2.50	13.50	20.00
6	3.00	15.50	12.90
7	3.50	18.80	17.55
8	4.00	22.70	17.18
9	4.50	25.40	10.63
10	5.00	27.10	6.27
11	5.50	28.20	3.90
12	6.00	29.00	2.76
13	6.50	29.70	2.36
14	7.00	30.30	1.98
15	7.50	30.80	1.62
16	8.00	31.30	1.60
17	8.50	31.70	1.26
18	9.00	32.20	1.55
19	9.50	32.70	1.53
20	10.00	33.10	1.21
21	10.50	33.50	1.19
22	11.00	33.80	0.89

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(33.8 - 33.76962) (1)/(1) = +3.04E-02 g/L Carbon
 (33.8 - 33.76962) (1)/(1) (12) = +2.53E-03 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3416

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2392DUP

Date: 11/27/98

Time: 23:04:43

Sample Size = 1 uL

Analyst : RD MEYERS

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 1.2 ug/minute C

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.10	63.64
3	1.50	4.80	77.08
4	2.00	14.50	66.90
5	2.50	24.60	41.06
6	3.00	31.90	22.88
7	3.50	36.40	12.36
8	4.00	39.00	6.67
9	4.50	40.40	3.47
10	5.00	41.10	1.70
11	5.50	41.50	0.96
12	6.00	41.90	0.95
13	6.50	42.20	0.71
14	7.00	42.40	0.47
15	7.50	42.70	0.70
16	8.00	43.00	0.70
17	8.50	43.20	0.46
18	9.00	43.40	0.46
19	9.50	43.60	0.46
20	10.00	43.90	0.68
21	10.50	44.10	0.45
22	11.00	44.30	0.45

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597

+1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(44.3 - 13.19648) (1)/(1) =

+3.11E+01 g/L Carbon

(44.3 - 13.19648) (1)/(1) (12) =

+2.59E+00 Molar Carbon

Sample Run By:

RD MEYERS

00002

3417

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2392DUP

Date: 11/27/98

Time: 23:18:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.20	0.00
2	1.01	1.40	85.71
3	1.50	6.80	79.41
4	2.00	11.30	39.82
5	2.50	14.00	19.29
6	3.00	16.50	15.15
7	3.50	20.70	20.29
8	4.00	24.00	13.75
9	4.50	26.30	8.75
10	5.00	27.60	4.71
11	5.50	28.50	3.16
12	6.00	29.40	3.06
13	6.50	30.00	2.00
14	7.00	30.60	1.96
15	7.50	31.10	1.61
16	8.00	31.60	1.58
17	8.50	32.00	1.25
18	9.00	32.50	1.54
19	9.50	32.90	1.22
20	10.00	33.40	1.50
21	10.50	33.80	1.18
22	11.00	34.10	0.88

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(34.1 - 33.75801) (1)/(1) = +3.42E-01 g/L Carbon
 (34.1 - 33.75801) (1)/(1) (12) = +2.85E-02 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3418

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 2822

Date: 11/27/98

Time: 23:34:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.51	18.60	95.16
4	2.00	139.20	86.64
5	2.50	313.70	55.63
6	3.00	475.10	33.97
7	3.50	597.30	20.46
8	4.00	678.50	11.97
9	4.50	725.80	6.52
10	5.00	751.40	3.41
11	5.50	763.60	1.60
12	6.00	769.50	0.77
13	6.50	772.40	0.38
14	7.00	774.30	0.25
15	7.50	775.50	0.15
16	8.00	776.50	0.13
17	8.50	777.40	0.12
18	9.00	778.20	0.10
19	9.50	778.90	0.09
20	10.00	779.60	0.09
21	10.50	780.30	0.09
22	11.00	781.00	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597

+1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(781 - 13.19648) (1)/(1) =

+7.678E+02 g/L Carbon

(781 - 13.19648) (1)/(1) (12) =

+6.398E+01 Molar Carbon

Sample Run By:

RD MEYERS

00002

3419

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2822

Date: 11/27/98

Time: 23:47:06

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.90	0.00
2	1.00	2.30	60.87
3	1.50	23.90	90.38
4	2.00	66.90	64.28
5	2.50	134.50	50.26
6	3.00	183.90	26.86
7	3.50	209.20	12.09
8	4.00	221.10	5.38
9	4.50	227.10	2.64
10	5.00	230.90	1.65
11	5.50	233.40	1.07
12	6.00	235.30	0.81
13	6.50	237.00	0.72
14	7.00	238.30	0.55
15	7.50	239.60	0.54
16	8.00	240.60	0.42
17	8.50	241.60	0.41
18	9.00	242.60	0.41
19	9.50	243.70	0.45
20	10.00	244.80	0.45
21	10.50	245.70	0.37
22	11.00	246.60	0.36

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(246.6 - 33.76063) (1)/(1) = +2.128E+02 g/L Carbon
 (246.6 - 33.76063) (1)/(1) (12) = +1.774E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3420

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 2822DUP

Date: 11/28/98

Time: 00:01:33

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst: RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	1.00	0.00
2	1.00	2.10	52.38
3	1.50	18.70	88.77
4	2.00	139.70	86.61
5	2.50	322.50	56.68
6	3.00	489.20	34.08
7	3.50	608.10	19.55
8	4.00	684.10	11.11
9	4.50	728.90	6.15
10	5.00	753.40	3.25
11	5.50	765.70	1.61
12	6.00	771.90	0.80
13	6.50	775.40	0.45
14	7.00	777.80	0.31
15	7.50	779.50	0.22
16	8.00	781.00	0.19
17	8.50	782.50	0.19
18	9.00	783.80	0.17
19	9.50	785.00	0.15
20	10.00	786.20	0.15
21	%-1429.50	787.50	0.17
22	%-1429.00	788.60	0.14

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 = +1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(788.6 --1714.803) (1)/(1) =

+2.503E+03 g/L Carbon

(788.6 --1714.803) (1)/(1) (12) =

+2.086E+02 Molar Carbon

Sample Run By:

RD MEYERS

00002

3421

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 2822DUP

Date: 11/28/98

Time: 00:15:57

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	1.70	0.00
2	1.00	3.30	48.48
3	1.50	22.20	85.14
4	2.00	64.30	65.47
5	2.50	132.10	51.32
6	3.00	182.40	27.58
7	3.50	208.50	12.52
8	4.00	220.70	5.53
9	4.50	227.60	3.03
10	5.00	231.80	1.81
11	5.50	235.10	1.40
12	6.00	237.60	1.05
13	6.50	239.80	0.92
14	7.00	241.90	0.87
15	7.50	243.70	0.74
16	8.00	245.50	0.73
17	8.50	247.00	0.61
18	9.00	248.50	0.60
19	9.50	249.90	0.56
20	10.00	251.40	0.60
21	10.50	252.80	0.55
22	11.00	254.20	0.55

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(254.2 - 33.76641) (1)/(1) = +2.204E+02 g/L Carbon
 (254.2 - 33.76641) (1)/(1) (12) = +1.837E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3422

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2823

Date: 11/28/98

Time: 00:31:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	1.90	0.00
2	1.00	3.50	45.71
3	1.50	12.90	72.87
4	2.00	92.70	86.08
5	2.50	224.20	58.65
6	3.00	348.20	35.61
7	3.50	437.80	20.47
8	4.00	494.30	11.43
9	4.50	526.30	6.08
10	5.00	543.50	3.16
11	5.50	552.10	1.56
12	6.00	556.70	0.83
13	6.50	560.90	0.75
14	7.00	564.80	0.69
15	7.50	567.40	0.46
16	8.00	569.50	0.37
17	8.50	571.40	0.33
18	9.00	573.20	0.31
19	9.50	574.90	0.30
20	10.00	576.50	0.28
21	10.50	578.10	0.28
22	11.00	579.80	0.29

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 =

+1.2E+00

ug/min Carbon

SAMPLE RESULTS:

(579.8 - 13.1974) (1)/(1) =

+5.666E+02 g/L Carbon

(579.8 - 13.1974) (1)/(1) (12) =

+4.722E+01 Molar Carbon

Sample Run By:

RD MEYERS

00002

3423

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV.2.0

Sample: 2823

Date: 11/28/98

Time: 00:44:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	1.90	0.00
2	1.01	3.70	48.65
3	1.50	18.80	80.32
4	2.00	55.00	65.82
5	2.50	121.40	54.70
6	3.00	178.40	31.95
7	3.50	209.90	15.01
8	4.00	225.50	6.92
9	4.50	233.60	3.47
10	5.00	238.50	2.05
11	5.50	242.20	1.53
12	6.00	245.10	1.18
13	6.50	247.70	1.05
14	7.00	250.10	0.96
15	7.50	252.10	0.79
16	8.00	254.00	0.75
17	8.50	256.00	0.78
18	9.00	257.70	0.66
19	9.50	259.40	0.66
20	10.00	261.10	0.65
21	10.50	262.60	0.57
22	11.00	264.50	0.72

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(264.5 - 33.76591) (1)/(1) = +2.307E+02 g/L Carbon
 (264.5 - 33.76591) (1)/(1) (12) = +1.923E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00002

3424

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2823DUP

Date: 11/28/98

Time: 00:58:20

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.2 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	2.30	0.00
2	1.01	4.40	47.73
3	1.51	15.80	72.15
4	2.00	108.30	85.41
5	2.50	245.30	55.85
6	3.00	363.00	32.42
7	3.50	443.40	18.13
8	4.00	493.40	10.13
9	4.50	521.40	5.37
10	5.00	535.80	2.69
11	5.50	543.20	1.36
12	6.00	547.40	0.77
13	6.50	550.20	0.51
14	7.00	552.30	0.38
15	7.50	554.30	0.36
16	8.00	556.10	0.32
17	8.50	557.80	0.30
18	9.00	559.50	0.30
19	9.50	561.10	0.29
20	10.00	562.80	0.30
21	10.50	564.30	0.27
22	11.00	565.80	0.27

USER INPUT BLANK VALUE

BLANK VALUE = 13.19517 micrograms carbon

BLANK FACTOR = 13.19517 / 10.99597 = +1.2E+00 ug/min Carbon

SAMPLE RESULTS:

(565.8 - 13.1986) (1)/(1) = +5.526E+02 g/L Carbon
 (565.8 - 13.1986) (1)/(1) (12) = +4.605E+01 Molar Carbon

Sample Run By:

RD MEYERS

00002

3425

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV: 2.0

Sample: 2823DUP

Date: 11/28/98

Time: 01:11:06

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	2.00	0.00
2	1.01	3.70	45.95
3	1.50	19.90	81.41
4	2.00	66.00	69.85
5	2.50	134.70	51.00
6	3.00	177.90	24.28
7	3.50	197.60	9.97
8	4.00	206.70	4.40
9	4.50	212.10	2.55
10	5.00	215.80	1.71
11	5.50	218.70	1.33
12	6.00	221.00	1.04
13	6.50	223.10	0.94
14	7.00	225.00	0.84
15	7.50	226.80	0.79
16	8.00	228.40	0.70
17	8.50	229.90	0.65
18	9.00	231.50	0.69
19	9.50	233.00	0.64
20	10.00	234.50	0.64
21	10.50	235.90	0.59
22	11.00	237.20	0.55

USER INPUT BLANK VALUE

BLANK VALUE = 33.75763 micrograms carbon

BLANK FACTOR = 33.75763 / 10.99597 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(237.2 - 33.76334) (1)/(1) = +2.034E+02 g/L Carbon
 (237.2 - 33.76334) (1)/(1) (12) = +1.695E+01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

RD MEYERS

00002

3426

LABCORE Completed Worklist Report for Worklist# 26745

Analyst: pjm

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@TICTOC1 TIC-02	LIQUID	6.02E+02	6.00E+2	99.668 % Recovery	
1 STD	0	@TICTOC1 TOC-02	LIQUID	3.00E+03	2.80E+3	93.333 % Recovery	
2 BLNK	0	@TICTOC1 TIC-02	LIQUID	1	7.00E-1	0.700 ug/mL	
2 BLNK	0	@TICTOC1 TOC-02	LIQUID	1	6.10E+0	6.100 ug/mL	
3 SAMPLE	S98T002824	0 @TICTOC1 TIC-02	LIQUID	N/A	5.29E+03	5.000 ug/mL	
3 SAMPLE	S98T002824	0 @TICTOC1 TOC-02	LIQUID	N/A	2.41E+03	40.000 ug/mL	
4 DUP	S98T002824	0 @TICTOC1 TIC-02	LIQUID	5.29E+3	5.23E+3	1.141 RPD	
4 DUP	S98T002824	0 @TICTOC1 TOC-02	LIQUID	2.41E+3	2.40E+3	0.416 RPD	
5 SAMPLE	S98T002825	0 @TICTOC1 TIC-02	LIQUID	N/A	6.06E+03	5.000 ug/mL	
5 SAMPLE	S98T002825	0 @TICTOC1 TOC-02	LIQUID	N/A	2.32E+03	40.000 ug/mL	
6 DUP	S98T002825	0 @TICTOC1 TIC-02	LIQUID	6.06E+3	6.17E+3	1.799 RPD	
6 DUP	S98T002825	0 @TICTOC1 TOC-02	LIQUID	2.32E+3	2.32E+3	0.000 RPD	
7 SAMPLE	S98T002826	0 @TICTOC1 TIC-02	LIQUID	N/A	3.72E+03	5.000 ug/mL	
7 SAMPLE	S98T002826	0 @TICTOC1 TOC-02	LIQUID	N/A	2.54E+03	40.000 ug/mL	
8 DUP	S98T002826	0 @TICTOC1 TIC-02	LIQUID	3.72E+3	3.88E+3	4.211 RPD	
8 DUP	S98T002826	0 @TICTOC1 TOC-02	LIQUID	2.54E+3	2.67E+3	4.990 RPD	

Final page for worklist# 26745

Analyst Signature _____ Date _____

C. J. Jones *12/1/98*
Analyst Signature _____ Date _____

[Signature] *12/2/98*
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 26745

Analyst: RLC Instrument: CARB2 Book# TIC 25N12D
Method: LA-342-100 Rev/Mod F-1 TOC 26N12F
Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	LIQUID		
2 BLNK			@TICTOC1	LIQUID		
3 SAMPLE	S98T002824 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002824 0		@TICTOC1	LIQUID		
5 SAMPLE	S98T002825 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
6 DUP	S98T002825 0		@TICTOC1	LIQUID		
7 SAMPLE	S98T002826 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
8 DUP	S98T002826 0		@TICTOC1	LIQUID		

Final page for worklist # 26745

RL McCann 11/28/98
Analyst Signature Date

A. Quinn 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	604.9	581.4
26745	μg of Carbon from Baseline (C2)	5.1	21
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000

Matrix:
LIQUIDBatch Number:
98005346Rerun:
0Sample Prep:
N/ASample #:
STDInstrument Code:
CARB2Prepared By:
CJOChemist:
MJLAnalyst:
PJMDate Complete:
12/01/98Analysis Date:
11/28/98Analysis Time:
04:00 PMSample Point:
BY-105QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)QC Found in $\mu\text{g}/\text{mL}$ = (C1 - C2) * DF / SSQC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if C1 < C2QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
04:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.00E+02	2.80E+03
BY-105	Percent Standard Recovery	99.6	93.4

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:		Date:	

STANDARD.WB1 REV 1.0

342100ML

3429

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	4.4	14.9
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code			
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98005346			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon	7.00E-01	6.10E+00
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>[Signature]</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3430

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	534.1	261.6
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005346			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002824			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	5.29E+03	2.41E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>N/A</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	527.6	260.5
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code	Known µg of C from Original Sample	5.29E+3	2.41E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005346			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002824			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	5.23E+03	2.40E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SAM5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	610.8	252.6
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code			
@TICTOC1			

Matrix
LIQUID
Batch Number
98005346
Rerun
0
Sample Prep
N/A
Sample #
S98T002825
Instrument Code
CARB2
Prepared By
CJO
Chemist
MJL
Analyst
PJM
Date Complete
12/01/98
Analysis Date
11/28/98
Analysis Time
04:00 PM
Sample Point
BY-105

$\mu\text{g of Carbon/mL} = (C1 - C2) * DF / SS$
 $\mu\text{g of Carbon/mL for TIC} = 5 \text{ if } C1 < C2$
 $\mu\text{g of Carbon/mL for TOC} = 40 \text{ if } C1 < C2$

	TIC	TOC
Method Detection Limit in µg/mL	5	40
µg of Carbon/mL	6.06E+03	2.32E+03

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>N/A</i>	Date:	

WORKBOOK PAGE: DUP6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	622.5	253.3
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code	Known µg of C from Original Sample	6.06E+3	2.32E+3
@TICTOC1			

Matrix
LIQUID
Batch Number
98005346
Rerun
0
Sample Prep
N/A
Sample #
S98T002825
Instrument Code
CARB2
Prepared By
CJO
Chemist
MJL
Analyst
PJM
Date Complete
12/01/98
Analysis Date
11/28/98
Analysis Time
04:00 PM
Sample Point
BY-105

$\mu\text{g of Carbon/mL} = (C1 - C2) * DF / SS$
 $\mu\text{g of Carbon/mL for TIC} = 5 \text{ if } C1 < C2$
 $\mu\text{g of Carbon/mL for TOC} = 40 \text{ if } C1 < C2$

	TIC	TOC
Method Detection Limit in µg/mL	5	40
µg of Carbon/mL	6.17E+03	2.32E+03

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:		Date:	

WORKBOOK PAGE: SAM7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	377.5	275.3
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005346			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002826			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/mL	3.72E+03	2.54E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>MA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3435

WORKBOOK PAGE: DUP8

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	393.4	288.2
26745	µg of Carbon from Baseline (C2)	5.1	21
Test Code	Known µg of C from Original Sample	3.72E+3	2.54E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005346			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002826			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/28/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	3.88E+03	2.67E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>MA</i>	Date:	

HNF-1663 REV. 0
TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/28/98

Time: 10:01:57

Sample Size = 1000 uL
Dil Factor = 1
Blank ID # =
Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	2.60	84.62
4	1.51	72.90	96.43
5	2.00	234.50	68.91
6	2.50	391.80	40.15
7	3.00	494.50	20.77
8	3.50	550.70	10.21
9	4.00	580.00	5.05
10	4.50	593.30	2.24
11	5.00	598.30	0.84
12	5.50	600.20	0.32
13	6.00	601.20	0.17
14	6.50	601.90	0.12
15	7.00	602.40	0.08
16	7.50	602.90	0.08
17	8.00	603.30	0.07
18	8.50	603.60	0.05
19	9.00	603.90	0.05
20	9.50	604.30	0.07
21	10.00	604.50	0.03
22	10.50	604.90	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 5.144043 micrograms carbon

BLANK FACTOR = 5.144043 / 10.49805 = +4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(604.9 - 5.144522) (1) / (1000) = +5.998E-01 g/L Carbon
(604.9 - 5.144522) (1) / (1000) (12) = +4.998E-02 Molar Carbon

Sample Run By: PJ McCown 11/28/98
PJ MCCOWN 00000

1.0mL 25 N12D

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3431 TO 3454

3437

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/28/98

Time: 10:14:30

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.50	0.30	100.00
3	1.01	0.70	57.14
4	1.50	3.80	81.58
5	2.00	26.00	85.38
6	2.50	96.40	73.03
7	3.00	236.20	59.19
8	3.50	407.20	41.99
9	4.00	505.40	19.43
10	4.50	546.90	7.59
11	5.00	565.20	3.24
12	5.50	572.00	1.19
13	6.00	575.00	0.52
14	6.50	576.80	0.31
15	7.00	577.60	0.14
16	7.50	578.50	0.16
17	8.00	579.20	0.12
18	8.50	579.80	0.10
19	9.00	580.20	0.07
20	9.50	580.70	0.09
21	10.00	581.00	0.05
22	10.50	581.40	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 20.99609 micrograms carbon

BLANK FACTOR = 20.99609 / 10.49805 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(581.4 - 20.99634) (1) / (200) = +2.802E+00 g/L Carbon
 (581.4 - 20.99634) (1) / (200) (12) = +2.335E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN 11/28/98

00000

.200mc 26N12F

3438

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/28/98

Time: 09:36:30

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE-2
 Blank Value = N/A

Analyst : PJ MCCOWN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	0.80	25.00
4	1.50	1.20	33.33
5	2.00	1.50	20.00
6	2.50	1.80	16.67
7	3.00	2.10	14.29
8	3.50	2.30	8.70
9	4.00	2.50	8.00
10	4.50	2.80	10.71
11	5.00	2.90	3.45
12	5.50	3.20	9.37
13	6.00	3.30	3.03
14	6.50	3.60	8.33
15	7.00	3.70	2.70
16	7.50	3.90	5.13
17	8.00	4.10	4.88
18	8.50	4.30	4.65
19	9.00	4.60	6.52
20	9.50	4.80	4.17
21	10.00	4.90	2.04
22	10.50	5.10	3.92

BLANK VALUE = 5.1 micrograms carbon

BLANK FACTOR = 5.1 / 10.49902 =

+4.9E-01

ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/28/98

Time: 09:49:12

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE-2
 Blank Value = N/A

Analyst : PJ MCCOWN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	0.80	75.00
4	1.50	2.10	61.90
5	2.00	4.60	54.35
6	2.50	7.90	41.77
7	3.00	11.70	32.48
8	3.50	13.80	15.22
9	4.00	15.60	11.54
10	4.50	16.50	5.45
11	5.00	17.30	4.62
12	5.50	17.90	3.35
13	6.00	18.30	2.19
14	6.50	18.70	2.14
15	7.00	19.10	2.09
16	7.50	19.40	1.55
17	8.00	19.60	1.02
18	8.50	19.90	1.51
19	9.00	20.10	1.00
20	9.50	20.50	1.95
21	10.00	20.70	0.97
22	10.50	21.00	1.43

BLANK VALUE = 21 micrograms carbon

BLANK FACTOR = 21 / 10.49805 =

+2.00E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 11/28/98

Time: 10:29:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	0.40	50.00
4	1.51	0.60	33.33
5	2.00	0.80	25.00
6	2.50	1.10	27.27
7	3.00	1.30	15.38
8	3.50	1.50	13.33
9	4.00	1.70	11.76
10	4.50	2.00	15.00
11	5.00	2.10	4.76
12	5.50	2.30	8.70
13	6.00	2.60	11.54
14	6.50	2.80	7.14
15	7.00	3.00	6.67
16	7.50	3.20	6.25
17	8.00	3.40	5.88
18	8.50	3.50	2.86
19	9.00	3.80	7.89
20	9.50	4.00	5.00
21	10.00	4.20	4.76
22	10.50	4.40	4.55

USER INPUT BLANK VALUE

BLANK VALUE = 5.144043 micrograms carbon

BLANK FACTOR = 5.144043 / 10.49805 = +4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(4.4 - 5.144522) (1)/(1) = < 5.00 E-3 g/L Carbon
 (4.4 - 5.144522) (1)/(1) (12) = < 4.17 E-4 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 11/28/98

Time: 10:41:57

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	0.50	60.00
4	1.50	1.70	70.59
5	2.00	3.70	54.05
6	2.50	6.10	39.34
7	3.00	8.30	26.51
8	3.50	9.80	15.31
9	4.00	10.90	10.09
10	4.50	11.40	4.39
11	5.00	11.90	4.20
12	5.50	12.30	3.25
13	6.00	12.70	3.15
14	6.50	13.00	2.31
15	7.00	13.30	2.26
16	7.50	13.50	1.48
17	8.00	13.70	1.46
18	8.50	14.00	2.14
19	9.00	14.20	1.41
20	9.50	14.40	1.39
21	10.00	14.70	2.04
22	10.50	14.90	1.34

USER INPUT BLANK VALUE

BLANK VALUE = 20.99609 micrograms carbon

BLANK FACTOR = 20.99609 / 10.49805 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(14.9 - 20.99805) (1)/(1) = < 5.00 E-3 g/L Carbon
(14.9 - 20.99805) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: *PJ*

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002824

Date: 11/28/98

Time: 10:54:01

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.60	50.00
4	1.50	14.10	95.74
5	2.00	72.70	80.61
6	2.50	173.30	58.05
7	3.00	300.40	42.31
8	3.50	412.10	27.11
9	4.00	474.30	13.11
10	4.50	505.10	6.10
11	5.00	519.20	2.72
12	5.50	525.30	1.16
13	6.00	527.70	0.45
14	6.50	529.20	0.28
15	7.00	530.20	0.19
16	7.50	531.00	0.15
17	8.00	531.90	0.17
18	8.50	532.20	0.06
19	9.00	532.70	0.09
20	9.50	533.20	0.09
21	10.00	533.60	0.07
22	10.50	534.10	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 5.144043 micrograms carbon

BLANK FACTOR = 5.144043 / 10.49805 =

+4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(534.1 - 5.144522) (1)/(1) =

+5.290E+02 g/L Carbon

(534.1 - 5.144522) (1)/(1) (12) =

+4.408E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

-100 mL

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002824

Date: 11/28/98

Time: 11:06:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.00	1.40	85.71
4	1.50	5.10	72.55
5	2.00	18.90	73.02
6	2.50	64.20	70.56
7	3.00	135.00	52.44
8	3.50	189.60	28.80
9	4.00	220.70	14.09
10	4.50	236.90	6.84
11	5.00	245.20	3.38
12	5.50	249.60	1.76
13	6.00	252.90	1.30
14	6.50	254.80	0.75
15	7.00	256.30	0.59
16	7.50	257.50	0.47
17	8.00	258.50	0.39
18	8.50	259.30	0.31
19	9.00	260.10	0.31
20	9.50	260.70	0.23
21	10.00	261.20	0.19
22	10.50	261.60	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 20.99609 micrograms Carbon

BLANK FACTOR = 20.99609 / 10.49805 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(261.6 - 20.99609) (1)/(1) = +2.406E+02 g/L Carbon
 (261.6 - 20.99609) (1)/(1) (12) = +2.005E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: RJL

PJ MCCOWN

00000

-100mL

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2824 DUP

Date: 11/28/98

Time: 11:19:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.00	4.20	95.24
4	1.50	60.90	93.10
5	2.00	180.40	66.24
6	2.50	319.90	43.61
7	3.04	419.60	23.76
8	3.54	474.40	11.55
9	4.04	502.30	5.55
10	4.54	514.70	2.41
11	5.04	519.20	0.87
12	5.54	521.30	0.40
13	6.04	522.60	0.25
14	6.54	523.70	0.21
15	7.04	524.20	0.10
16	7.53	524.90	0.13
17	8.03	525.40	0.10
18	8.53	525.90	0.10
19	9.03	526.30	0.08
20	9.53	526.70	0.08
21	10.03	527.10	0.08
22	10.53	527.60	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 5.144043 micrograms carbon

BLANK FACTOR = 5.144043 / 10.49805 = +4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(527.6 - 5.160701) (1)/(1) = +5.224E+02 g/L Carbon
 (527.6 - 5.160701) (1)/(1)(12) = +4.354E+01 Molar Carbon

Sample Run By: PJ

PJ MCCOWN

00000

.100mL

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2824 DUP

Date: 11/28/98

Time: 11:30:44

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.80	62.50
4	1.51	4.30	81.40
5	2.00	19.60	78.06
6	2.50	69.70	71.88
7	3.00	142.30	51.02
8	3.50	194.60	26.88
9	4.00	223.10	12.77
10	4.50	238.50	6.46
11	5.00	245.70	2.93
12	5.50	249.70	1.60
13	6.00	252.10	0.95
14	6.50	254.00	0.75
15	7.00	255.50	0.59
16	7.50	256.70	0.47
17	8.00	257.70	0.39
18	8.50	258.30	0.23
19	9.00	259.10	0.31
20	9.50	259.60	0.19
21	10.00	260.10	0.19
22	10.50	260.50	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 20.99609 micrograms carbon

BLANK FACTOR = 20.99609 / 10.49805 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(260.5 - 20.99634) (1) / (1) = +2.395E+02 g/L Carbon
 (260.5 - 20.99634) (1) / (1) (12) = +1.996E+01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002825

Date: 11/28/98

Time: 11:42:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.90	66.67
4	1.51	36.00	97.50
5	2.00	162.80	77.89
6	2.50	323.60	49.69
7	3.00	450.80	28.22
8	3.50	527.30	14.51
9	4.00	569.40	7.39
10	4.50	591.10	3.67
11	5.00	600.40	1.55
12	5.50	603.90	0.58
13	6.00	605.50	0.26
14	6.50	606.50	0.16
15	7.00	607.30	0.13
16	7.50	608.10	0.13
17	8.00	608.60	0.08
18	8.50	609.30	0.11
19	9.00	609.50	0.03
20	9.50	610.00	0.08
21	10.00	610.40	0.07
22	10.50	610.80	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 5.144043 micrograms carbon

BLANK FACTOR = 5.144043 / 10.49805 = +4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(610.8 - 5.144522) (1) / (1) = +6.057E+02 g/L Carbon
 (610.8 - 5.144522) (1) / (1) (12) = +5.047E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002825

Date: 11/28/98

Time: 11:55:55

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	2.20	0.00
2	1.01	14.00	84.29
3	1.51	52.70	73.43
4	2.01	121.90	56.77
5	2.50	177.90	31.48
6	3.00	210.80	15.61
7	3.50	228.00	7.54
8	4.00	236.50	3.59
9	4.50	241.00	1.87
10	5.00	243.50	1.03
11	5.50	245.30	0.73
12	6.00	246.40	0.45
13	6.50	247.40	0.40
14	7.00	248.20	0.32
15	7.50	248.80	0.24
16	8.00	249.50	0.28
17	8.50	250.40	0.36
18	9.00	250.90	0.20
19	9.50	251.40	0.20
20	10.00	251.80	0.16
21	10.50	252.20	0.16
22	11.00	252.60	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(252.6 - 21.99768) (1)/(1) =

+2.306E+02 g/L Carbon

(252.6 - 21.99768) (1)/(1) (12) =

+1.922E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: *PJ McCown*

PJ MCCOWN

00000

.100 mL

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2825 DUP

Date: 11/28/98

Time: 12:08:27

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.70	76.47
3	1.51	50.80	96.65
4	2.00	190.70	73.36
5	2.50	355.10	46.30
6	3.00	476.30	25.45
7	3.50	546.90	12.91
8	4.00	585.80	6.64
9	4.50	604.60	3.11
10	5.00	612.20	1.24
11	5.50	615.20	0.49
12	6.00	616.90	0.28
13	6.50	617.90	0.16
14	7.00	618.60	0.11
15	7.50	619.40	0.13
16	8.00	619.80	0.06
17	8.50	620.40	0.10
18	9.00	620.80	0.06
19	9.50	621.30	0.08
20	10.00	621.80	0.08
21	10.50	622.20	0.06
22	11.00	622.50	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(622.5 - 5.389432) (1)/(1) =
(622.5 - 5.389432) (1)/(1) (12) =

+6.171E+02 g/L Carbon
+5.143E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2825 DUP

Date: 11/28/98

Time: 13:25:51

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.00	60.00
3	1.51	3.80	73.68
4	2.00	16.20	76.54
5	2.50	56.50	71.33
6	3.00	123.40	54.21
7	3.50	177.90	30.64
8	4.00	210.40	15.45
9	4.50	227.60	7.56
10	5.00	236.20	3.64
11	5.50	240.40	1.75
12	6.00	243.00	1.07
13	6.50	244.80	0.74
14	7.00	246.80	0.81
15	7.50	248.40	0.64
16	8.00	249.50	0.44
17	8.50	250.50	0.40
18	9.00	251.10	0.24
19	9.50	251.80	0.28
20	10.00	252.30	0.20
21	10.50	252.70	0.16
22	11.00	253.30	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(253.3 - 21.99609) (1)/(1) =

+2.313E+02 g/L Carbon

(253.3 - 21.99609) (1)/(1)(12) =

+1.928E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ McC

PJ MCCOWN

00000

.100 mL

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002826

Date: 11/28/98

Time: 13:42:42

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.60	50.00
3	1.51	8.20	92.68
4	2.01	49.00	83.27
5	2.51	117.20	58.19
6	3.00	200.10	41.43
7	3.50	280.00	28.54
8	4.00	328.70	14.82
9	4.50	353.90	7.12
10	5.00	365.00	3.04
11	5.50	369.20	1.14
12	6.00	371.20	0.54
13	6.50	372.50	0.35
14	7.00	373.30	0.21
15	7.50	374.10	0.21
16	8.00	374.70	0.16
17	8.50	375.20	0.13
18	9.00	375.80	0.16
19	9.50	376.20	0.11
20	10.00	376.70	0.13
21	10.50	377.10	0.11
22	11.00	377.50	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(377.5 - 5.389462) (1)/(1) =

+3.721E+02 g/L Carbon

(377.5 - 5.389462) (1)/(1) (12) =

+3.101E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002826

Date: 11/28/98

Time: 13:55:27

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.10	54.55
3	1.50	4.40	75.00
4	2.00	17.90	75.42
5	2.50	64.40	72.20
6	3.00	139.80	53.93
7	3.50	198.50	29.57
8	4.00	232.90	14.77
9	4.50	251.00	7.21
10	5.00	259.40	3.24
11	5.50	263.70	1.63
12	6.00	265.90	0.83
13	6.50	267.90	0.75
14	7.00	269.20	0.48
15	7.50	270.50	0.48
16	8.00	271.50	0.37
17	8.50	272.30	0.29
18	9.00	273.00	0.26
19	9.50	273.70	0.26
20	10.00	274.40	0.26
21	10.50	275.00	0.22
22	11.00	275.30	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(275.3 - 21.99768) (1)/(1) =

+2.533E+02 g/L Carbon

(275.3 - 21.99768) (1)/(1) (12) =

+2.111E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

.100 m -

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2826 DUP

Date: 11/28/98

Time: 14:07:38

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # =
 Blank Value = .49 ug/minute C

Analyst : PJ MCCOWN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.51	8.90	89.89
4	2.00	59.20	84.97
5	2.50	144.90	59.14
6	3.04	225.20	35.66
7	3.54	288.10	21.83
8	4.04	335.90	14.23
9	4.54	362.40	7.31
10	5.04	376.10	3.64
11	5.54	382.90	1.78
12	6.04	385.60	0.70
13	6.54	387.20	0.41
14	7.04	388.30	0.28
15	7.53	389.30	0.26
16	8.03	390.00	0.18
17	8.53	390.70	0.18
18	9.03	391.20	0.13
19	9.53	391.80	0.15
20	10.03	392.40	0.15
21	10.53	392.90	0.13
22	11.03	393.40	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+4.9E-01 ug/min Carbon

SAMPLE RESULTS:

(393.4 - 5.406001) (1)/(1) =

+3.880E+02 g/L Carbon

(393.4 - 5.406001) (1)/(1) (12) =

+3.233E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.100 mL

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2826 DUP

Date: 11/28/98

Time: 14:20:34

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2 ug/minute C

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.00	1.40	57.14
3	1.50	6.90	79.71
4	2.00	31.10	77.81
5	2.50	94.60	67.12
6	3.00	169.50	44.19
7	3.50	219.80	22.88
8	4.00	248.50	11.55
9	4.50	263.40	5.66
10	5.00	270.80	2.73
11	5.50	275.30	1.63
12	6.00	278.10	1.01
13	6.50	280.00	0.68
14	7.00	281.40	0.50
15	7.50	283.10	0.60
16	8.00	284.10	0.35
17	8.50	285.00	0.32
18	9.00	285.80	0.28
19	9.50	286.50	0.24
20	10.00	287.10	0.21
21	10.50	287.70	0.21
22	11.00	288.20	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 0 micrograms carbon

BLANK FACTOR = 0 / 0 =

+2.0E+00

ug/min Carbon

SAMPLE RESULTS:

(288.2 - 21.99548) (1)/(1) =

+2.662E+02 g/L Carbon

(288.2 - 21.99548) (1)/(1) (12) =

+2.218E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ McCown

PJ MCCOWN

00000

LABCORE Completed Worklist Report for Worklist# 26746

Analyst: pjm

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	LIQUID	6.02E+02	6.05E+2	100.498 % Recovery	
1 STD	0		@TICTOC1 TOC-02	LIQUID	3.00E+03	2.84E+3	94.667 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	LIQUID	1	8.00E-1	0.800 ug/mL	
2 BLNK	0		@TICTOC1 TOC-02	LIQUID	1	8.70E+0	8.700 ug/mL	
3 SAMPLE	S98T002827	0	@TICTOC1 TIC-02	LIQUID	N/A	3.28E+03	5.000 ug/mL	
3 SAMPLE	S98T002827	0	@TICTOC1 TOC-02	LIQUID	N/A	2.76E+03	40.000 ug/mL	
4 DUP	S98T002827	0	@TICTOC1 TIC-02	LIQUID	3.28E+3	3.29E+3	0.304 RPD	
4 DUP	S98T002827	0	@TICTOC1 TOC-02	LIQUID	2.76E+3	2.74E+3	0.727 RPD	
5 SPK	S98T002827	0	@TICTOC1 TIC-02	LIQUID	1.00E+02	1.03E+02	103.000 % Recovery	
5 SPK	S98T002827	0	@TICTOC1 TOC-02	LIQUID	1.00E+02	9.73E+01	97.300 % Recovery	
6 SAMPLE	S98T002828	0	@TICTOC1 TIC-02	LIQUID	N/A	3.14E+03	5.000 ug/mL	
6 SAMPLE	S98T002828	0	@TICTOC1 TOC-02	LIQUID	N/A	2.67E+03	40.000 ug/mL	
7 DUP	S98T002828	0	@TICTOC1 TIC-02	LIQUID	3.14E+3	3.23E+3	2.826 RPD	
7 DUP	S98T002828	0	@TICTOC1 TOC-02	LIQUID	2.67E+3	2.80E+3	4.753 RPD	
8 SAMPLE	S98T002829	0	@TICTOC1 TIC-02	LIQUID	N/A	2.72E+03	5.000 ug/mL	
8 SAMPLE	S98T002829	0	@TICTOC1 TOC-02	LIQUID	N/A	2.53E+3	40.000 ug/mL	
9 DUP	S98T002829	0	@TICTOC1 TIC-02	LIQUID	2.72E+3	2.85E+3	4.668 RPD	
9 DUP	S98T002829	0	@TICTOC1 TOC-02	LIQUID	2.53E+3	2.64E+3	4.255 RPD	

Final page for worklist# 26746

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 26746

Analyst: RKL Instrument: CARB2 _____ Book# TIC 25N12D
Method: LA-342-100 Rev/Mod F-1 TOC 26N12F

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	LIQUID		
2 BLNK			@TICTOC1	LIQUID		
3 SAMPLE	S98T002827 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002827 0		@TICTOC1	LIQUID		
5 SPK	S98T002827 0		@TICTOC1	LIQUID		
6 SAMPLE	S98T002828 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T002828 0		@TICTOC1	LIQUID		
8 SAMPLE	S98T002829 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
9 DUP	S98T002829 0		@TICTOC1	LIQUID		

Final page for worklist # 26746

B. McCown 11/27/98
Analyst Signature Date

C. J. Paulin 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	614.5	601.5
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000

Matrix:
LIQUIDBatch Number
98005347Return
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
CJOChemist
MJLAnalyst
PJMDate Complete
12/01/98Analysis Date
11/27/98Analysis Time
04:00 PMSample Point
BY-105

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.05E+02	2.84E+03
BY-105	Percent Standard Recovery	100.4	94.6

Data Entered By:

Signature of Chemist:

CJO

Date: 12/01/98

Date:

STANDARD.WB1 REV 1.0

342100ML

3457

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	9.2	25.2
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon	8.00E-01	8.70E+00
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:		Date:	
BLANK.WB1 REV 1.0	342100ML		

3458

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	338.2	310.3
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002827			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	3.28E+03	2.76E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>MA</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	339.4	307.8
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code	Known µg of C from Original Sample	3.28E+3	2.76E+3
@TICTOC1			

Matrix
LIQUID
Batch Number
98005347
Rerun
0
Sample Prep
N/A
Sample #
S98T002827
Instrument Code
CARB2
Prepared By
CJO
Chemist
MJL
Analyst
PJM
Date Complete
12/01/98
Analysis Date
11/27/98
Analysis Time
04:00 PM
Sample Point
BY-105

$\mu\text{g of Carbon/mL} = (C1 - C2) * DF / SS$
 $\mu\text{g of Carbon/mL for TIC} = 5 \text{ if } C1 < C2$
 $\mu\text{g of Carbon/mL for TOC} = 40 \text{ if } C1 < C2$

	TIC	TOC
Method Detection Limit in µg/mL	5	40
µg of Carbon/mL	3.29E+03	2.74E+03

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:		Date:	

SAMPLE.WB1 REV 1.0

342100ML

3460

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Volume in mL (SS)	0.1000	0.1000
Work List	Final Coulometer Reading in µg (C1)	338.2	310.3
26746	Spiked Vial Data		
Test Code	Sample Volume in mL (SPK SS)	0.1000	0.1000
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	647.3	602.3
LIQUID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/ml (SPK CONC)	602	3000
98005347	µg C in baseline (BL)	10	33.9
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002827			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time			
04:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.18E+02	2.92E+03
BY-105	Percent Spike Recovery	102.7	97.3

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>N/A</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3461

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	324.3	301.3
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002828			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	3.14E+03	2.67E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:		Date:	

SAMPLE.WB1 REV 1.0

342100ML

3462

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	332.6	314.1
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code	Known µg of C from Original Sample	3.14E+3	2.67E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002828			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	3.23E+03	2.80E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	NA	Date:	

WORKBOOK PAGE: SAM8

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	282.2	286.8
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002829			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	2.72E+03	2.53E+03
BY-105			

Data Entered By:

CJO

Date:

12/01/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3464

I:\342100\OUT\26746.WB1

12/01/98

13:39:32

WORKBOOK PAGE: DUP9

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	294.6	298.2
26746	µg of Carbon from Baseline (C2)	10	33.9
Test Code	Known µg of C from Original Sample	2.72E+3	2.53E+3
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98005347			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002829			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 PM			
Sample Point	µg of Carbon/mL	2.85E+03	2.64E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/27/98

Time: 10:15:40

Sample Size = 1 uL
Dil Factor = 1
Blank ID # = BASE-2
Blank Value = N/A

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.00	0.80	62.50
4	1.50	1.20	33.33
5	2.00	1.80	33.33
6	2.50	2.30	21.74
7	3.00	2.70	14.81
8	3.50	3.20	15.63
9	4.00	3.70	13.51
10	4.50	4.20	11.90
11	5.00	4.70	10.64
12	5.50	5.10	7.84
13	6.00	5.60	8.93
14	6.50	6.10	8.20
15	7.00	6.50	6.15
16	7.50	7.10	8.45
17	8.00	7.60	6.58
18	8.50	8.10	6.17
19	9.00	8.60	5.81
20	9.50	9.00	4.44
21	10.00	9.40	4.26
22	10.50	10.00	6.00

BLANK VALUE = 10 micrograms carbon

BLANK FACTOR = 10. / 10.49811 =

+9.53E-01 ug/min Carbon

Sample Run By: PJ McCown

00000

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3466 TO 3485.

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV.2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/27/98

Time: 10:28:13

Sample Size = 1 uL
Dil Factor = 1
Blank ID # = BASE-2
Blank Value = N/A

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

==	Reading	====	Analysis Time	====	Coulometer	====	% Difference	==
1			0.08		0.00		0.00	
2			0.51		0.40		100.00	
3			1.01		0.90		55.56	
4			1.50		1.50		40.00	
5			2.00		3.90		61.54	
6			2.50		8.80		55.68	
7			3.00		13.90		36.69	
8			3.50		18.90		26.46	
9			4.00		22.00		14.09	
10			4.50		24.20		9.09	
11			5.00		25.60		5.47	
12			5.50		26.90		4.83	
13			6.00		27.70		2.89	
14			6.50		28.60		3.15	
15			7.00		29.40		2.72	
16			7.50		30.30		2.97	
17			8.00		31.00		2.26	
18			8.50		31.50		1.59	
19			9.00		32.10		1.87	
20			9.50		32.70		1.83	
21			10.00		33.30		1.80	
22			10.50		33.90		1.77	

BLANK VALUE = 33.9 micrograms carbon

BLANK FACTOR = 33.9 / 10.49799 = +3.23E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/27/98

Time: 10:45:54

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.00	100.00
3	1.01	1.70	41.18
4	1.50	2.20	22.73
5	2.00	6.20	64.52
6	2.50	45.30	86.31
7	3.00	141.00	67.87
8	3.50	259.00	45.56
9	4.00	373.20	30.60
10	4.50	467.00	20.09
11	5.00	529.80	11.85
12	5.50	566.90	6.54
13	6.00	587.60	3.52
14	6.50	598.70	1.85
15	7.00	604.20	0.91
16	7.50	607.70	0.58
17	8.00	609.60	0.31
18	8.50	611.00	0.23
19	9.00	612.10	0.18
20	9.50	613.10	0.16
21	10.00	613.80	0.11
22	10.50	614.50	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(614.5 - 10.00557) (1)/(1000) = +6.045E-01 g/L Carbon
 (614.5 - 10.00557) (1)/(1000) (12) = +5.037E-02 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

1.0 mL 25 NITR

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: STD-2

Date: 11/27/98

Time: 10:57:59

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	1.00	50.00
4	1.51	1.60	37.50
5	2.00	3.20	50.00
6	2.50	9.40	65.96
7	3.00	33.50	71.94
8	3.50	98.20	65.89
9	4.00	213.50	54.00
10	4.50	347.10	38.49
11	5.00	454.80	23.68
12	5.50	519.90	12.52
13	6.00	557.50	6.74
14	6.50	577.30	3.43
15	7.00	587.10	1.67
16	7.50	592.10	0.84
17	8.00	595.20	0.52
18	8.50	597.00	0.30
19	9.00	598.30	0.22
20	9.50	599.40	0.18
21	10.00	600.50	0.18
22	10.50	601.50	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(601.5 - 33.91185) (1)/(200) = +2.838E+00 g/L Carbon
 (601.5 - 33.91185) (1)/(200) (12) = +2.365E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

-200ML 26W12 F

3469

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTUC REV 2.0

Sample: BLNK-2

Date: 11/27/98

Time: 11:17:45

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis	Time	==== Coulometer	==== % Difference ==
1	0.08		0.00	0.00
2	0.51		0.50	100.00
3	1.01		1.00	50.00
4	1.50		1.40	28.57
5	2.00		2.00	30.00
6	2.50		2.30	13.04
7	3.00		2.80	17.86
8	3.50		3.10	9.68
9	4.00		3.60	13.89
10	4.50		4.20	14.29
11	5.00		4.60	8.70
12	5.50		5.00	8.00
13	6.00		5.40	7.41
14	6.50		5.80	6.90
15	7.00		6.20	6.45
16	7.50		6.70	7.46
17	8.00		7.10	5.63
18	8.50		7.50	5.33
19	9.00		8.00	6.25
20	9.50		8.60	6.98
21	10.03		8.90	3.37
22	10.53		9.20	3.26

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(9.2 - 10.03686) (1)/(1) = < 5.00 E-3 g/L Carbon
 (9.2 - 10.03686) (1)/(1) (12) = < 4.17 E-4 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV.2.0

Sample: BLNK-2

Date: 11/27/98

Time: 11:32:37

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 3.23 ug/minute C

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	0.80	50.00
4	1.50	1.30	38.46
5	2.00	2.60	50.00
6	2.50	5.50	52.73
7	3.00	8.90	38.20
8	3.50	12.40	28.23
9	4.00	15.00	17.33
10	4.50	17.20	12.79
11	5.00	18.30	6.01
12	5.50	19.40	5.67
13	6.00	20.10	3.48
14	6.50	20.80	3.37
15	7.00	21.50	3.26
16	7.50	22.10	2.71
17	8.00	22.70	2.64
18	8.50	23.20	2.16
19	9.00	23.80	2.52
20	9.50	24.20	1.65
21	10.00	24.70	2.02
22	10.50	25.20	1.98

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(25.2 - 33.90908) (1)/(1) = < 5.00 E-3 g/L Carbon

(25.2 - 33.90908) (1)/(1) (12) = < 4.17 E-4 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002827

Date: 11/27/98

Time: 11:49:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	0.80	50.00
4	1.51	1.40	42.86
5	2.00	3.60	61.11
6	2.50	15.70	77.07
7	3.00	50.90	69.16
8	3.50	98.20	48.17
9	4.00	148.60	33.92
10	4.50	200.40	25.85
11	5.00	248.40	19.32
12	5.50	284.90	12.81
13	6.00	307.50	7.35
14	6.50	319.60	3.79
15	7.00	326.60	2.14
16	7.50	330.40	1.15
17	8.00	333.00	0.78
18	8.50	334.40	0.42
19	9.00	335.60	0.36
20	9.50	336.60	0.30
21	10.00	337.50	0.27
22	10.50	338.20	0.21

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(338.2 - 10.00476) (1) / (1) = +3.282E+02 g/L Carbon
 (338.2 - 10.00476) (1) / (1) (12) = +2.735E+01 Molar Carbon

Sample Run By: PJ McCown

PJ MCCOWN

00000

.100 mL

HNF-1663 REV. 0

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002827

Date: 11/27/98

Time: 12:01:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	1.20	50.00
4	1.50	1.80	33.33
5	2.01	3.90	53.85
6	2.50	11.90	67.23
7	3.00	34.90	65.90
8	3.50	88.70	60.65
9	4.00	158.00	43.86
10	4.50	215.90	26.82
11	5.00	253.70	14.90
12	5.50	275.10	7.78
13	6.00	287.00	4.15
14	6.50	294.10	2.41
15	7.00	298.60	1.51
16	7.50	301.60	0.99
17	8.00	303.70	0.69
18	8.50	305.70	0.65
19	9.00	306.90	0.39
20	9.50	308.10	0.39
21	10.00	309.40	0.42
22	10.50	310.30	0.29

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(310.3 - 33.91185) (1)/(1) = +2.764E+02 g/L Carbon
 (310.3 - 33.91185) (1)/(1)(12) = +2.303E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ
PJ MCCOWN

00000

.100mL

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2827 DUP

Date: 11/27/98

Time: 12:13:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	1.70	52.94
4	1.51	2.30	26.09
5	2.00	3.70	37.84
6	2.50	13.50	72.59
7	3.00	41.60	67.55
8	3.50	84.10	50.54
9	4.00	132.50	36.53
10	4.50	193.10	31.38
11	5.00	249.50	22.61
12	5.50	287.50	13.22
13	6.00	309.50	7.11
14	6.50	321.80	3.82
15	7.00	328.00	1.89
16	7.50	331.70	1.12
17	8.00	334.00	0.69
18	8.50	335.60	0.48
19	9.00	336.90	0.39
20	9.50	337.70	0.24
21	10.00	338.60	0.27
22	10.50	339.40	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(339.4 - 10.00476) (1)/(1) = +3.294E+02 g/L Carbon
 (339.4 - 10.00476) (1)/(1)(12) = +2.745E+01 Molar Carbon

Sample Run By: *PJM*

PJ MCCOWN

00000

•100MC

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2827 DUP

Date: 11/27/98

Time: 12:26:18

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.00	1.00	50.00
4	1.50	1.70	41.18
5	2.00	3.70	54.05
6	2.50	13.00	71.54
7	3.00	43.90	70.39
8	3.50	105.80	58.51
9	4.00	175.70	39.78
10	4.50	226.50	22.43
11	5.00	259.20	12.62
12	5.50	277.50	6.59
13	6.00	287.90	3.61
14	6.50	294.10	2.11
15	7.00	297.70	1.21
16	7.50	300.30	0.87
17	8.00	302.30	0.66
18	8.50	303.80	0.49
19	9.00	305.00	0.39
20	9.50	306.20	0.39
21	10.00	307.00	0.26
22	10.50	307.80	0.26

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(307.8 - 33.90849) (1)/(1) = +2.739E+02 g/L Carbon
 (307.8 - 33.90849) (1)/(1) (12) = +2.282E+01 Molar Carbon

<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

PJ MCCOWN

00000

.100 mL

3475

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0Sample: 2728 ^{ML 11/30/98} SPK

Date: 11/27/98

Time: 12:41:37

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.40	50.00
4	1.50	2.20	36.36
5	2.00	15.20	85.53
6	2.50	81.30	81.30
7	3.00	196.70	58.67
8	3.50	326.30	39.72
9	4.00	436.50	25.25
10	4.50	517.70	15.68
11	5.00	572.00	9.49
12	5.50	605.00	5.45
13	6.00	622.50	2.81
14	6.50	631.60	1.44
15	7.00	637.00	0.85
16	7.50	639.70	0.42
17	8.00	641.70	0.31
18	8.50	643.40	0.26
19	9.00	644.60	0.19
20	9.50	645.50	0.14
21	10.00	646.50	0.15
22	10.50	647.30	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(647.3 - 10.00557) (1)/(1) = +6.373E+02 g/L Carbon
 (647.3 - 10.00557) (1)/(1) (12) = +5.311E+01 Molar Carbon

Sample Run By: PJ McCown

PJ MCCOWN

00000

.100mL

+ .500mL 25W12D

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2827 SPK

Date: 11/27/98

Time: 12:55:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	1.20	33.33
4	1.50	1.80	33.33
5	2.00	5.40	66.67
6	2.50	20.90	74.16
7	3.00	71.20	70.65
8	3.50	180.40	60.53
9	4.00	320.50	43.71
10	4.50	433.30	26.03
11	5.00	503.50	13.94
12	5.50	545.00	7.61
13	6.00	567.70	4.00
14	6.50	580.50	2.20
15	7.00	587.60	1.21
16	7.50	591.90	0.73
17	8.00	595.00	0.52
18	8.50	596.80	0.30
19	9.00	598.60	0.30
20	9.50	600.10	0.25
21	10.00	601.10	0.17
22	10.50	602.30	0.20

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(602.3 - 33.90849) (1)/(1) = +5.684E+02 g/L Carbon
 (602.3 - 33.90849) (1)/(1) (12) = +4.737E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: *PJM*

PJ MCCOWN

00000

*100mL

+100mL 26N12F

3477

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002828

Date: 11/27/98

Time: 14:45:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.60	50.00
4	1.51	1.00	40.00
5	2.04	1.80	44.44
6	2.54	6.40	71.88
7	3.04	21.90	70.78
8	3.54	50.10	56.29
9	4.04	84.70	40.85
10	4.54	124.00	31.69
11	5.04	179.30	30.84
12	5.54	234.40	23.51
13	6.04	273.30	14.23
14	6.54	295.90	7.64
15	7.04	308.20	3.99
16	7.54	314.80	2.10
17	8.04	318.20	1.07
18	8.54	320.10	0.59
19	9.04	321.60	0.47
20	9.54	322.50	0.28
21	10.04	323.60	0.34
22	10.54	324.30	0.22

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 =

+9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(324.3 - 10.04053) (1)/(1) =

+3.143E+02 g/L Carbon

(324.3 - 10.04053) (1)/(1)(12) =

+2.619E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002828

Date: 11/27/98

Time: 14:58:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	0.90	44.44
4	1.50	1.60	43.75
5	2.00	3.30	51.52
6	2.50	10.00	67.00
7	3.00	31.30	68.05
8	3.50	81.80	61.74
9	4.00	148.80	45.03
10	4.50	205.80	27.70
11	5.00	244.80	15.93
12	5.50	267.10	8.35
13	6.00	279.30	4.37
14	6.50	286.60	2.55
15	7.00	290.90	1.48
16	7.50	294.00	1.05
17	8.00	296.00	0.68
18	8.50	297.60	0.54
19	9.00	299.00	0.47
20	9.50	299.80	0.27
21	10.00	300.70	0.30
22	10.50	301.30	0.20

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(301.3 - 33.90908) (1)/(1) = +2.674E+02 g/L Carbon
 (301.3 - 33.90908) (1)/(1) (12) = +2.228E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2828 DUP

Date: 11/27/98

Time: 15:10:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	1.00	50.00
4	1.50	1.40	28.57
5	2.00	2.40	41.67
6	2.50	9.00	73.33
7	3.00	35.90	74.93
8	3.50	83.20	56.85
9	4.00	139.60	40.40
10	4.50	193.50	27.86
11	5.00	241.90	20.01
12	5.50	279.80	13.55
13	6.00	303.20	7.72
14	6.50	315.40	3.87
15	7.00	321.80	1.99
16	7.50	325.90	1.26
17	8.00	327.80	0.58
18	8.50	329.30	0.46
19	9.00	330.50	0.36
20	9.50	331.30	0.24
21	10.00	332.00	0.21
22	10.50	332.60	0.18

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 =

+9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(332.6 - 10.00458) (1)/(1) =

+3.226E+02 g/L Carbon

(332.6 - 10.00458) (1)/(1) (12) =

+2.688E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2828 DUP

Date: 11/27/98

Time: 15:22:10

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	1.20	50.00
4	1.51	1.90	36.84
5	2.00	5.60	66.07
6	2.50	20.10	72.14
7	3.00	59.60	66.28
8	3.50	123.70	51.82
9	4.00	188.00	34.20
10	4.50	236.40	20.47
11	5.00	267.30	11.56
12	5.50	285.20	6.28
13	6.00	295.20	3.39
14	6.50	300.80	1.86
15	7.00	304.40	1.18
16	7.50	307.10	0.88
17	8.00	309.10	0.65
18	8.50	310.30	0.39
19	9.00	311.40	0.35
20	9.50	312.50	0.35
21	10.00	313.30	0.26
22	10.50	314.10	0.25

USER INPUT BLANK VALUE :

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(314.1 - 33.915) (1)/(1) = +2.802E+02 g/L Carbon
 (314.1 - 33.915) (1)/(1) (12) = +2.335E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: *PJ*

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T002829

Date: 11/27/98

Time: 15:35:41

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.50	0.70	100.00
3	1.00	1.30	46.15
4	1.50	1.90	31.58
5	2.00	3.60	47.22
6	2.50	16.20	77.78
7	3.00	56.10	71.12
8	3.50	110.30	49.14
9	4.00	165.10	33.19
10	4.50	209.50	21.19
11	5.00	240.20	12.78
12	5.50	257.70	6.79
13	6.00	267.00	3.48
14	6.50	272.60	2.05
15	7.00	275.40	1.02
16	7.50	277.30	0.69
17	8.00	278.40	0.40
18	8.50	279.40	0.36
19	9.00	280.20	0.29
20	9.50	280.90	0.25
21	10.00	281.50	0.21
22	10.50	282.20	0.25

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 =

+9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(282.2 - 10.00458) (1)/(1) =

+2.722E+02 g/L Carbon

(282.2 - 10.00458) (1)/(1) (12) =

+2.268E+01 Molar Carbon

Sample Run By: PJ

PJ MCCOWN

00000

*100mL

3482

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002829

Date: 11/27/98

Time: 15:48:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.00	30.00
4	1.50	1.60	37.50
5	2.00	3.80	57.89
6	2.50	12.30	69.11
7	3.00	33.10	62.84
8	3.50	76.80	56.90
9	4.00	135.40	43.28
10	4.50	188.70	28.25
11	5.00	227.20	16.95
12	5.50	250.40	9.27
13	6.00	263.50	4.97
14	6.50	270.70	2.66
15	7.00	276.00	1.92
16	7.50	278.60	0.93
17	8.00	281.10	0.89
18	8.50	282.70	0.57
19	9.00	284.00	0.46
20	9.50	285.10	0.39
21	10.00	286.10	0.35
22	10.50	286.80	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(286.8 - 33.91165) (1)/(1) = +2.529E+02 g/L Carbon
 (286.8 - 33.91165) (1)/(1) (12) = +2.107E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN
 PJ MCCOWN

00000

HNF-1663 REV. 0

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2829 DUP

Date: 11/27/98

Time: 16:00:07

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .953 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	1.10	45.45
4	1.50	1.70	35.29
5	2.00	2.90	41.38
6	2.50	7.90	63.29
7	3.00	26.50	70.19
8	3.50	57.20	53.67
9	4.00	91.90	37.76
10	4.50	124.30	26.07
11	5.00	154.00	19.29
12	5.50	193.10	20.25
13	6.00	232.70	17.02
14	6.50	260.30	10.60
15	7.00	275.70	5.59
16	7.50	283.40	2.72
17	8.00	288.10	1.63
18	8.50	290.50	0.83
19	9.00	291.90	0.48
20	9.50	292.90	0.34
21	10.00	293.80	0.31
22	10.50	294.60	0.27

USER INPUT BLANK VALUE

BLANK VALUE = 10.00458 micrograms carbon

BLANK FACTOR = 10.00458 / 10.49799 = +9.5E-01 ug/min Carbon

SAMPLE RESULTS:

(294.6 - 10.00557) (1)/(1) =

+2.846E+02 g/L Carbon

(294.6 - 10.00557) (1)/(1) (12) =

+2.372E+01 Molar Carbon

Sample Run By: PJ McCown

PJ MCCOWN

00000

• 100mL

3484

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2829 DUP

Date: 11/27/98

Time: 16:12:59

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.23 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.20	41.67
4	1.50	1.70	29.41
5	2.00	3.60	52.78
6	2.50	12.40	70.97
7	3.00	36.90	66.40
8	3.50	87.10	57.63
9	4.00	148.80	41.47
10	4.50	201.60	26.19
11	5.00	238.40	15.44
12	5.50	261.70	8.90
13	6.00	274.60	4.70
14	6.50	282.20	2.69
15	7.00	287.00	1.67
16	7.50	290.00	1.03
17	8.00	292.10	0.72
18	8.50	293.80	0.58
19	9.00	295.40	0.54
20	9.50	296.50	0.37
21	10.00	297.50	0.34
22	10.50	298.20	0.23

USER INPUT BLANK VALUE

BLANK VALUE = 33.90849 micrograms carbon

BLANK FACTOR = 33.90849 / 10.49799 = +3.2E+00 ug/min Carbon

SAMPLE RESULTS:

(298.2 - 33.90869) (1)/(1) = +2.643E+02 g/L Carbon
 (298.2 - 33.90869) (1)/(1) (12) = +2.202E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ

PJ MCCOWN

00000

3485

LABCORE Completed Worklist Report for Worklist# 26747

Analyst: pjm Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@TICTOC1 TIC-02	LIQUID	6.02E+02	6.07E+2	100.831 % Recovery	
1 STD	0	@TICTOC1 TOC-02	LIQUID	3.00E+03	2.76E+3	92.000 % Recovery	
2 BLNK	0	@TICTOC1 TIC-02	LIQUID	1	7.00E-1	0.700 ug/mL	
2 BLNK	0	@TICTOC1 TOC-02	LIQUID	1	9.60E+0	9.600 ug/mL	
3 SAMPLE	S98T002940 0	@TICTOC1 TIC-02	LIQUID	N/A	3.11E+03	5.000 ug/mL	
3 SAMPLE	S98T002940 0	@TICTOC1 TOC-02	LIQUID	N/A	2.72E+03	40.000 ug/mL	
4 DUP	S98T002940 0	@TICTOC1 TIC-02	LIQUID	3.11E+3	3.04E+3	2.276 RPD	
4 DUP	S98T002940 0	@TICTOC1 TOC-02	LIQUID	2.72E+3	2.68E+3	1.481 RPD	
5 SPK	S98T002940 0	@TICTOC1 TIC-02	LIQUID	1.00E+02	1.04E+02	104.000 % Recovery	
5 SPK	S98T002940 0	@TICTOC1 TOC-02	LIQUID	1.00E+02	1.01E+02	101.000 % Recovery	
6 SAMPLE	S98T002965 0	@TICTOC1 TIC-02	LIQUID	N/A	7.26E+03	5.000 ug/mL	
6 SAMPLE	S98T002965 0	@TICTOC1 TOC-02	LIQUID	N/A	2.18E+03	40.000 ug/mL	
7 DUP	S98T002965 0	@TICTOC1 TIC-02	LIQUID	7.26E+3	7.05E+3	2.935 RPD	
7 DUP	S98T002965 0	@TICTOC1 TOC-02	LIQUID	2.18E+3	2.16E+3	0.922 RPD	
8 SPK	S98T002965 0	@TICTOC1 TIC-02	LIQUID	1.00E+02	1.12E+02	112.000 % Recovery	
8 SPK	S98T002965 0	@TICTOC1 TOC-02	LIQUID	1.00E+02	9.81E+01	98.100 % Recovery	

Final page for worklist# 26747

Analyst Signature _____ Date _____

Belmont Sant 11-13-98
Analyst Signature _____ Date _____

[Signature] 11/17/98
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 26747

Analyst: RLM Instrument: CARB2 _____ Book# TIC 25N12D
Method: LA-342-100 Rev/Mod F-1 TOC 26N12F

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	LIQUID		
2 BLNK			@TICTOC1	LIQUID		
3 SAMPLE	S98T002940 0		@TICTOC1	LIQUID	98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002940 0		@TICTOC1	LIQUID		
5 SPK	S98T002940 0		@TICTOC1	LIQUID		
6 SAMPLE	S98T002965 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T002965 0		@TICTOC1	LIQUID		
8 SPK	S98T002965 0		@TICTOC1	LIQUID		
9 SAMPLE	S98T002966 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
10 DUP	S98T002966 0		@TICTOC1	LIQUID		

*Did not run
out of time*

Final page for worklist # 26747

R. McCann 11/12/98
Analyst Signature Date

_____
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	612.1	580.7
26747	μg of Carbon from Baseline (C2)	5.2	29
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000

Matrix
LIQUIDBatch Number
98005348Rerun
0Sample Prep
N/ASample #
STANDARDInstrument Code
CARB2Prepared By
JMVChemist
MJLAnalyst
PJMDate Complete
11/13/98Analysis Date
11/12/98Analysis Time
02:12 PMSample Point
BY-105QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)QC Found in $\mu\text{g}/\text{mL}$ = (C1 - C2) * DF / SSQC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if C1 < C2QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
02:12 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.07E+02	2.76E+03
BY-105	Percent Standard Recovery	100.8	92.0

Data Entered By:	JMV	Date:	11/13/98
Signature of Chemist:		Date:	

STANDARD.WB1 REV 1.0

342100ML

3488

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	5.9	19.4
26747	µg of Carbon from Baseline (C2)	5.2	29
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005348			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
02:12 PM			
Sample Point	µg of Carbon	7.00E-01	9.60E+00
BY-105			

Data Entered By:	JMV	Date:	11/13/98
Signature of Chemist:	<i>NK</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3489

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	316.1	300.5
26747	µg of Carbon from Baseline (C2)	5.2	29
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005348			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002940			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
02:12 PM			
Sample Point	µg of Carbon/mL	3.11E+03	2.72E+03
BY-105			

Data Entered By:

JMV

Date:

11/13/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3490

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	309.6	297.3
26747	µg of Carbon from Baseline (C2)	5.2	29
Test Code	Known µg of C from Original Sample	3.11E+3	2.72E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005348			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002940			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
02:12 PM			
Sample Point	µg of Carbon/mL	3.04E+03	2.68E+03
BY-105			

Data Entered By:

JM V

Date:

11/13/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3491

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Volume in mL (SS)	0.1000	0.1000
Work List	Final Coulometer Reading in µg (C1)	316.1	300.5
26747	Spiked Vial Data		
Test Code	Sample Volume in mL (SPK SS)	0.1000	0.1000
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	628.2	602.5
LIQUID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/ml (SPK CONC)	602	3000
98005348	µg C in baseline (BL)	5.2	29
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002940			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time			
02:12 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.24E+02	3.02E+03
BY-105	Percent Spike Recovery	103.7	100.7

Data Entered By:	JMV	Date:	11/13/98
Signature of Chemist:		Date:	

SPIKE.WB1 REV 1.3

342100ML

3492

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	731.6	246.5
26747	µg of Carbon from Baseline (C2)	5.2	29
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005348			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002965			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
02:12 PM			
Sample Point	µg of Carbon/mL	7.26E+03	2.18E+03
BY-105			

Data Entered By:

JMV

Date: 11/13/98

Signature of Chemist:

A/A

Date:

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	710.1	244.9
26747	µg of Carbon from Baseline (C2)	5.2	29
Test Code	Known µg of C from Original Sample	7.26E+3	2.18E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005348			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002965			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/13/98			
Analysis Date			
11/12/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
02:12 PM			
Sample Point	µg of Carbon/mL	7.05E+03	2.16E+03
BY-105			

Data Entered By:	JMV	Date:	11/13/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3494

WORKBOOK PAGE: SPIKE8

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Volume in mL (SS)	0.1000	0.1000
Work List	Final Coulometer Reading in µg (C1)	731.6	246.5
26747	Spiked Vial Data		
Test Code	Sample Volume in mL (SPK SS)	0.1000	0.1000
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	1069.2	540.7
LIQUID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98005348	µg C in baseline (BL)	5.2	29

Rerun	
0	
Sample Prep	
N/A	
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100
S98T002965	
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)
CARB2	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100
Prepared By	
JMV	
Chemist	
MJL	
Analyst	
PJM	
Date Complete	
11/13/98	
Analysis Date	
11/12/98	
Analysis Time	
02:12 PM	
Sample Point	
BY-105	

	TIC	TOC
QC Actual in µg/mL	6.02E+02	3.00E+03
QC Found in µg/mL	6.75E+02	2.94E+03
Percent Spike Recovery	112.2	98.1

Data Entered By:	JMV	Date:	11/13/98
Signature of Chemist:	NA	Date:	

SPIKE.WB1 REV 1.3

342100ML

3495

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/12/98

Time: 09:13:22

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Concentration	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	6.40	89.71
4	1.51	120.90	94.38
5	2.00	297.50	59.33
6	2.50	438.30	32.17
7	3.00	525.20	16.55
8	3.50	570.10	7.88
9	4.00	521.60	3.63
10	4.50	600.50	1.48
11	5.00	604.50	0.66
12	5.50	606.30	0.30
13	6.00	607.70	0.23
14	6.50	608.50	0.13
15	7.00	609.10	0.10
16	7.50	609.60	0.08
17	8.00	610.00	0.07
18	8.50	610.60	0.10
19	9.00	611.00	0.07
20	9.50	611.40	0.07
21	10.00	611.80	0.07
22	10.50	612.10	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(612.1 - 5.249512) (1) / (1000) = +6.069E-01 g/L Carbon
 (612.1 - 5.249512) (1) / (1000) (12) = +5.057E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3496 TO 3513.Sample Run By: PJ McCown 11/12/98 00000
PJ MCCOWN

1.0 mL 25N121) 3496

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/12/98

Time: 09:26:02

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.76 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	3.00	0.00
2	0.51	3.10	100.00
3	1.00	2.10	80.95
4	1.50	15.80	86.71
5	2.00	75.50	79.07
6	2.50	216.70	65.16
7	3.00	389.00	45.69
8	3.50	504.00	20.83
9	4.00	545.20	7.56
10	4.50	561.30	2.87
11	5.00	563.30	1.23
12	5.50	571.80	0.61
13	6.00	573.50	0.30
14	6.50	574.80	0.23
15	7.00	575.90	0.19
16	7.50	576.80	0.16
17	8.00	577.30	0.14
18	8.50	578.30	0.12
19	9.00	579.10	0.14
20	9.50	579.60	0.09
21	10.00	580.10	0.09
22	10.50	580.70	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(580.7 - 28.97714) (1) / (200) = +2.759E+00 g/L Carbon
 (580.7 - 28.97714) (1) / (200) (12) = +2.299E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.3 ug/min Carbon!!!!>>>>

Sample Run By:

PJ McCown 11/12/98

PJ MCCOWN

00000

.200mc 26N12 F

3497

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/12/98

Time: 08:44:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-2

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.00	100.00
3	1.01	0.00	50.00
4	1.51	0.70	42.86
5	2.00	1.00	30.00
6	2.50	1.10	9.09
7	3.00	1.50	26.67
8	3.50	1.80	16.67
9	4.00	1.80	0.00
10	4.50	2.20	18.18
11	5.00	2.30	4.35
12	5.50	2.60	17.86
13	6.00	3.80	0.00
14	6.50	3.20	12.50
15	7.00	2.40	5.88
16	7.50	3.60	5.56
17	8.00	3.90	7.69
18	8.50	4.10	4.88
19	9.00	4.50	8.89
20	9.50	4.60	2.17
21	10.00	4.90	6.12
22	10.50	5.20	5.77

BLANK VALUE = 5.2 micrograms carbon

BLANK FACTOR = 5.2 / 10.49817 =

+5.0E-01

ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/12/98

Time: 08:58:44

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-2

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Colomometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	1.90	84.21
4	1.51	5.30	64.15
5	2.00	11.00	51.82
6	2.50	16.30	32.52
7	3.00	19.90	18.09
8	3.50	21.80	8.72
9	4.00	22.80	4.39
10	4.50	23.80	4.20
11	5.00	24.40	2.46
12	5.50	25.10	2.79
13	6.00	25.60	1.95
14	6.50	25.80	0.78
15	7.00	26.20	1.53
16	7.50	26.60	1.50
17	8.00	27.10	1.85
18	8.50	27.50	1.45
19	9.00	27.80	1.08
20	9.50	28.10	1.07
21	10.00	28.40	1.06
22	10.50	29.00	2.07

BLANK VALUE = 29 micrograms carbon

BLANK FACTOR = 29 / 10.49915 =

+2.76E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 3.0

Sample: BLNK-2

Date: 11/12/98

Time: 09:42:14

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.00	100.00
3	1.01	0.00	62.50
4	1.51	1.00	20.00
5	2.00	1.50	23.08
6	2.50	1.50	13.33
7	3.00	1.80	16.67
8	3.50	2.10	14.29
9	4.00	2.50	16.00
10	4.50	2.60	3.85
11	5.00	3.00	13.33
12	5.50	3.30	9.09
13	6.00	3.60	8.33
14	6.50	3.90	7.69
15	7.00	4.00	2.50
16	7.50	4.20	4.76
17	8.00	4.50	6.67
18	8.50	4.80	6.25
19	9.00	5.10	5.88
20	9.50	5.40	5.56
21	10.00	5.60	3.57
22	10.50	5.90	5.08

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(5.9 - 5.249481) (1)/(1) = +6.5E-01 g/L Carbon
 (5.9 - 5.249481) (1)/(1) (12) = +5.4E-02 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 3.0

Sample: BLNK-2

Date: 11/12/98

Time: 09:59:29

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.76 ug/minute

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.10	100.00
3	1.01	0.40	75.00
4	1.51	1.70	76.47
5	2.00	4.70	63.83
6	2.50	8.30	43.37
7	3.00	11.10	25.23
8	3.50	13.00	14.62
9	4.00	14.10	7.80
10	4.50	14.90	5.37
11	5.00	15.50	3.87
12	5.50	16.00	3.13
13	6.00	16.50	3.03
14	6.50	16.90	2.37
15	7.00	17.30	2.31
16	7.50	17.60	1.70
17	8.00	17.80	1.12
18	8.50	18.20	2.20
19	9.00	18.50	1.62
20	9.50	18.80	1.60
21	10.00	19.10	1.57
22	10.50	19.40	1.55

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(19.4 - 28.9773) (1)/(1) = < 5.00 E-3 g/L Carbon
 (19.4 - 28.9773) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

 PJ MCCOWN

00000

3501

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002940

Date: 11/12/98

Time: 10:44:04

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Colloimeter	==== % Difference ==
1	0.08	0.00	0.00
2	0.50	0.10	100.00
3	1.00	1.60	75.00
4	1.50	11.40	85.96
5	2.00	49.90	77.15
6	2.50	96.70	49.44
7	3.00	143.90	31.41
8	3.50	214.40	32.88
9	4.00	269.20	20.56
10	4.50	292.70	7.79
11	5.00	302.10	3.11
12	5.50	306.30	1.37
13	6.00	308.70	0.78
14	6.50	310.10	0.45
15	7.00	311.20	0.39
16	7.50	312.30	0.32
17	8.00	312.90	0.19
18	8.50	313.80	0.29
19	9.00	314.50	0.22
20	9.50	315.00	0.16
21	10.00	315.70	0.22
22	10.50	316.10	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(316.1 - 5.249054) (1)/(1) = +3.109E+02 g/L Carbon
 (316.1 - 5.249054) (1)/(1)(12) = +2.590E+01 Molar Carbon

Sample Run By: *PJ McCown*

PJ MCCOWN

00000

.100ML

3502

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002940

Date: 11/12/98

Time: 12:01:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.76 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	3.10	88.24
4	1.51	3.40	83.22
5	2.00	102.60	72.26
6	2.50	195.50	43.94
7	3.00	242.80	19.48
8	3.50	265.70	8.62
9	4.00	277.90	4.39
10	4.50	264.50	2.32
11	5.00	288.10	1.25
12	5.50	290.80	0.93
13	6.00	293.00	0.75
14	6.50	294.20	0.41
15	7.00	295.20	0.34
16	7.50	296.00	0.27
17	8.00	296.90	0.30
18	8.50	297.50	0.34
19	9.00	298.90	0.33
20	9.50	299.30	0.13
21	10.00	300.00	0.23
22	10.50	300.50	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(300.5 - 28.9773) (1)/(1) = +2.715E+02 g/L Carbon
 (300.5 - 28.9773) (1)/(12) = +2.263E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.3 ug/min Carbon!!!!>>>>

Sample Run By: RJm

PJ MCCOWN

00000

1.00 mL

3503

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2940 DUP

Date: 11/12/93

Time: 12:19:27

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.54	0.40	100.00
3	1.04	1.50	73.33
4	1.54	15.00	90.00
5	2.04	58.40	74.32
6	2.54	104.50	44.11
7	3.04	144.40	27.63
8	3.54	198.70	27.33
9	4.04	256.10	22.41
10	4.54	282.50	9.35
11	5.04	293.40	3.72
12	5.54	298.60	1.74
13	6.04	301.10	0.83
14	6.54	302.90	0.59
15	7.04	304.10	0.39
16	7.54	305.10	0.33
17	8.04	306.00	0.29
18	8.54	306.70	0.23
19	9.03	307.50	0.26
20	9.54	308.10	0.19
21	10.03	308.90	0.26
22	10.53	309.60	0.23

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(309.6 - 5.266846) (1) / (1) = +3.043E+02 g/L Carbon
 (309.6 - 5.266846) (1) / (1) (12) = +2.536E+01 Molar Carbon

Sample Run By: PJ

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2940 ^{DUP Rm} ~~SPK~~ 11/12/98 Date: 11/12/93 Time: 12:34:58

Sample Size = 1 uL Analyst : PJ MCCOWN
Dil Factor = 1 Min Readings = 22
Blank ID # = Max Readings = 22
Blank Value = 2.76 ug/minute C % Difference = 10

== Reading	==== Analysis Time	==== Calometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.00	0.90	55.56
4	1.51	4.30	79.07
5	2.00	19.20	77.60
6	2.50	71.40	73.11
7	3.00	155.90	54.20
8	3.50	216.30	27.92
9	4.00	249.40	13.27
10	4.50	267.10	6.63
11	5.00	276.60	3.43
12	5.50	282.50	2.09
13	6.00	286.00	1.22
14	6.50	288.50	0.87
15	7.00	290.10	0.55
16	7.50	292.00	0.65
17	8.00	293.30	0.44
18	8.50	294.30	0.34
19	9.00	295.20	0.30
20	9.50	295.90	0.24
21	10.00	296.50	0.20
22	10.50	297.30	0.27

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(297.3 - 28.97444) (1) / (1) = +2.683E+02 g/L Carbon
 (297.3 - 28.97444) (1) / (1) (12) = +2.236E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: Rm

PJ MCCOWN

00000

.100m - 3505

4.100m - 26.112 ^{Rm} 11/12/98

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2 0

Sample: 2940 SPK

Date: 11/12/98

Time: 12:49:59

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.30	100.00
3	1.00	2.10	38.10
4	1.50	43.10	95.13
5	2.00	202.30	78.70
6	2.50	263.60	44.36
7	3.00	475.60	23.55
8	3.50	543.90	12.56
9	4.00	581.90	6.53
10	4.50	601.10	3.19
11	5.00	610.40	1.52
12	5.50	618.40	0.81
13	6.00	618.40	0.49
14	6.50	626.20	0.29
15	7.00	621.90	0.27
16	7.50	623.10	0.19
17	8.00	624.20	0.18
18	8.50	625.10	0.14
19	9.00	626.10	0.16
20	9.50	626.80	0.11
21	10.00	627.50	0.11
22	10.50	628.20	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(628.2 - 5.248993) (1)/(1) = +6.230E+02 g/L Carbon
 (628.2 - 5.248993) (1)/(1) (12) = +5.191E+01 Molar Carbon

Sample Run By: PJ

PJ MCCOWN

00000

.100
 + .500 25N12D

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2940 SPK

Date: 11/12/98

Time: 13:02:10

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.76 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	3.40	76.47
4	1.50	21.40	84.11
5	2.00	80.00	73.25
6	2.50	220.30	63.69
7	3.00	391.50	43.73
8	3.50	494.70	20.86
9	4.00	542.00	8.73
10	4.50	565.90	4.22
11	5.00	578.00	2.09
12	5.50	584.50	1.11
13	6.00	588.50	0.68
14	6.50	591.30	0.47
15	7.00	593.20	0.32
16	7.50	595.50	0.39
17	8.00	597.10	0.27
18	8.50	598.50	0.23
19	9.00	599.60	0.18
20	9.50	600.80	0.20
21	10.00	601.50	0.13
22	10.50	602.50	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(602.5 - 28.97495) (1)/(1) = +5.735E+02 g/L Carbon
 (602.5 - 28.97495) (1)/(1)(12) = +4.779E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

 PJ MCCOWN

00000

-100 mL

3507

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002965

Date: 11/12/98

Time: 13:16:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.00	3.70	91.95
4	1.51	151.10	94.24
5	2.00	370.50	59.22
6	2.50	513.50	31.83
7	3.00	639.80	15.05
8	3.50	685.90	6.72
9	4.00	706.80	2.96
10	4.50	715.60	1.23
11	5.00	720.00	0.61
12	5.50	722.50	0.35
13	6.00	724.30	0.25
14	6.50	725.40	0.15
15	7.00	726.40	0.14
16	7.50	727.40	0.14
17	8.00	728.20	0.11
18	8.50	729.10	0.12
19	9.00	729.70	0.08
20	9.50	730.40	0.10
21	10.00	731.00	0.08
22	10.50	731.60	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(731.6 - 5.249512) (1) / (1) = +7.264E+02 g/L Carbon
 (731.6 - 5.249512) (1) / (1) (12) = +6.053E+01 Molar Carbon

Sample Run By: PJ

PJ MCCOWN

00000

.100mL

3508

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002965

Date: 11/12/98

Time: 13:30:05

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.76 ug/minute C

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	2.70	74.07
4	1.50	10.50	74.29
5	2.00	31.10	66.24
6	2.50	88.40	64.82
7	3.00	152.10	41.88
8	3.50	191.70	20.66
9	4.00	212.40	9.75
10	4.50	221.20	4.84
11	5.00	230.00	2.96
12	5.50	234.70	2.00
13	6.00	237.10	1.01
14	6.50	238.80	0.71
15	7.00	240.10	0.54
16	7.50	241.30	0.50
17	8.00	242.60	0.54
18	8.50	243.70	0.45
19	9.00	244.30	0.25
20	9.50	245.20	0.37
21	10.00	246.00	0.33
22	10.50	246.50	0.20

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(246.5 - 28.97495) (1) / (1) = +2.175E+02 g/L Carbon
 (246.5 - 28.97495) (1) / (1) (12) = +1.813E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2 0

Sample: 2965 DUP

Date: 11/12/98

Time: 13:42:35

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .5 ug/minute C

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.60	0.00
2	0.51	10.60	100.00
3	1.01	8.60	93.02
4	1.50	163.00	94.72
5	2.00	385.90	57.76
6	2.50	543.50	29.00
7	3.00	628.20	13.48
8	3.50	669.30	6.14
9	4.00	688.00	2.72
10	4.50	695.50	1.08
11	5.00	699.10	0.51
12	5.50	701.30	0.31
13	6.00	702.80	0.21
14	6.50	704.10	0.18
15	7.00	705.00	0.13
16	7.50	705.90	0.13
17	8.00	706.80	0.13
18	8.50	707.70	0.13
19	9.00	708.10	0.06
20	9.50	708.90	0.11
21	10.00	709.50	0.08
22	10.50	710.10	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 =

+5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(710.1 - 5.249481) (1)/(1) =

+7.049E+02 g/L Carbon

(710.1 - 5.249481) (1)/(1) (12) =

+5.874E+01 Molar Carbon

Sample Run By: *PJ*

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 2965 DUP

Date: 11/12/98

Time: 13:56:09

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.76 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.50	0.70	100.00
3	1.01	3.70	81.08
4	1.50	16.80	77.98
5	2.04	55.10	69.51
6	2.54	124.00	55.56
7	3.04	176.00	29.55
8	3.54	204.10	13.77
9	4.04	218.70	6.68
10	4.54	226.40	3.40
11	5.04	230.80	1.91
12	5.54	234.00	1.37
13	6.04	235.80	0.76
14	6.54	237.20	0.59
15	7.04	238.60	0.59
16	7.54	240.00	0.58
17	8.03	241.10	0.46
18	8.53	241.80	0.29
19	9.03	242.90	0.45
20	9.53	243.50	0.25
21	10.03	244.50	0.41
22	10.53	244.90	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.42915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(244.9 - 29.06827) (1)/(1) = +2.158E+02 g/L Carbon
 (244.9 - 29.06827) (1)/(1) (12) = +1.799E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

.100 mC

3511

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: 2965 SPK

Date: 11/12/98

Time: 14:12:05

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .5 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.50	100.00
3	1.01	135.90	98.90
4	1.50	471.50	71.20
5	2.00	751.80	37.23
6	2.50	911.80	17.55
7	3.00	989.40	7.84
8	3.50	1029.10	3.86
9	4.00	1046.80	1.69
10	4.50	1054.10	0.69
11	5.00	1057.60	0.33
12	5.50	1059.80	0.21
13	6.00	1061.50	0.16
14	6.50	1062.60	0.10
15	7.00	1063.60	0.09
16	7.50	1064.70	0.10
17	8.00	1065.60	0.08
18	8.50	1066.40	0.08
19	9.00	1067.20	0.07
20	9.50	1067.90	0.07
21	10.00	1068.60	0.07
22	10.50	1069.20	0.06

USER INPUT BLANK VALUE

BLANK VALUE = 5.249573 micrograms carbon

BLANK FACTOR = 5.249573 / 10.49915 = +5.0E-01 ug/min Carbon

SAMPLE RESULTS:

(1069.2 - 5.249054) (1)/(1) =

+1.0640E+03 g/L Carbon

(1069.2 - 5.249054) (1)/(1) (12) =

+8.8663E+01 Molar Carbon

Sample Run By: *PJ*

PJ MCCOWN

00000

.100mL
+ .500mL 25N/12 3512

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: 2965 SPK

Date: 11/12/98

Time: 14:26:06

Sample Size = 1 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.76 ug/minute C

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	2.40	66.67
4	1.51	10.90	77.98
5	2.00	36.30	69.97
6	2.51	120.00	69.75
7	3.01	272.70	56.00
8	3.51	404.80	32.63
9	4.00	472.00	14.24
10	4.50	504.00	6.35
11	5.00	518.90	2.87
12	5.50	525.70	1.29
13	6.00	529.70	0.76
14	6.50	532.00	0.43
15	7.00	534.20	0.41
16	7.50	535.50	0.24
17	8.00	536.60	0.20
18	8.50	537.70	0.20
19	9.00	538.80	0.20
20	9.50	539.50	0.13
21	10.00	540.20	0.13
22	10.50	540.70	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 28.97764 micrograms carbon

BLANK FACTOR = 28.97764 / 10.49915 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(540.7 - 28.9773) (1)/(1) = +5.117E+02 g/L Carbon
 (540.7 - 28.9773) (1)/(1) (12) = +4.264E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN

PJ MCCOWN

00000

.100
 +.100 26N12F
 3513

LABCORE Completed Worklist Report for Worklist# 26749

Analyst: slh

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	5.02E+02	5.00E+2	99.668 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.76E+3	92.000 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	1.40E+0	1.400 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	1.07E+1	10.700 ug/g	
3 SAMPLE	S98T002589	0	@TICTOC1 TIC-02	SOLID	N/A	5.06E+03	5.000 ug/g	
3 SAMPLE	S98T002589	0	@TICTOC1 TOC-02	SOLID	N/A	4.00E+01	40.000 ug/g	
4 DUP	S98T002589	0	@TICTOC1 TIC-02	SOLID	5.06E+3	3.20E+3	45.036 RPD	
4 DUP	S98T002589	0	@TICTOC1 TOC-02	SOLID	<4.00E+1	<4.00E+1	RPD	
5 TRIPL	S98T002589	0	@TICTOC1 TIC-02	SOLID	5.06E+3	3.75E+3	29.739 RPD	
5 TRIPL	S98T002589	0	@TICTOC1 TOC-02	SOLID	<4.00E+1	<4.00E+1	RPD	
6 SPK	S98T002589	0	@TICTOC1 TIC-02	SOLID	1.00E+02	1.01E+02	101.000 % Recovery	
6 SPK	S98T002589	0	@TICTOC1 TOC-02	SOLID	1.00E+02	9.68E+01	96.800 % Recovery	
7 SAMPLE	S98T002590	0	@TICTOC1 TIC-02	SOLID	N/A	1.50E+04	5.000 ug/g	
7 SAMPLE	S98T002590	0	@TICTOC1 TOC-02	SOLID	N/A	5.83E+01	40.000 ug/g	
8 DUP	S98T002590	0	@TICTOC1 TIC-02	SOLID	1.50E+4	1.29E+4	15.054 RPD	
8 DUP	S98T002590	0	@TICTOC1 TOC-02	SOLID	5.83E+1	9.59E+1	48.768 RPD	
9 SAMPLE	S98T002591	0	@TICTOC1 TIC-02	SOLID	N/A	3.98E+04	5.000 ug/g	
9 SAMPLE	S98T002591	0	@TICTOC1 TOC-02	SOLID	N/A	3.76E+03	40.000 ug/g	
10 DUP	S98T002591	0	@TICTOC1 TIC-02	SOLID	3.98E+4	4.39E+4	9.797 RPD	
10 DUP	S98T002591	0	@TICTOC1 TOC-02	SOLID	3.76E+3	3.90E+3	3.655 RPD	

Final page for worklist# 26749

Analyst Signature _____ Date _____

Analyst Signature *C. J. Quinn* 11/25/98
Date

Reviewer Signature *[Signature]* 12/1/98
Date

10/23/98 09:37
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26749

Analyst: SLH Instrument: CARB2Book# TU 25N12-DMethod: LA-342-100 Rev/Mod F.1TUC 26N12-FWorklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	SOLID		
2 BLNK			@TICTOC1	SOLID		
3 SAMPLE	S98T002589 0		@TICTOC1	SOLID	98000436	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002589 0		@TICTOC1	SOLID		
5 SPK	S98T002589 0		@TICTOC1	SOLID		
6 SAMPLE	S98T002590 0		@TICTOC1	SOLID	98000436	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T002590 0		@TICTOC1	SOLID		
8 SAMPLE	S98T002591 0		@TICTOC1	SOLID	98000436	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
9 DUP	S98T002591 0		@TICTOC1	SOLID		

Final page for worklist # 26749

Sandra H. Hood Boetger
 Analyst Signature Date
 11-21-98

Ch. Quinn 11/25/98
 Analyst Signature Date

Filed out
11-22-98
SLH

Data Entry Comments: Batch 5350

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	10.7	26.6
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:30 PM		5	40
Sample Point	µg of Carbon	1.40E+00	1.07E+01
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3516

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	609.1	588.6
26749	μg of Carbon from Baseline (C2)	9.3	37.3
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000

Matrix
LIQUIDBatch Number
98005350Rerun
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
CJOChemist
MJLAnalyst
SLHDate Complete
11/25/98Analysis Date
11/21/98Analysis Time
05:30 PMSample Point
BY-105QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
05:30 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.00E+02	2.76E+03
BY-105	Percent Standard Recovery	99.6	91.9

Data Entered By:

Signature of Chemist:

CJO

Date: 11/25/98

Date:

STANDARD.WB1 REV 1.0

342100ML

3517

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0580	0.0580
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	302.5	32.8
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002589			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date	NOTE: FOR TOC: The Reported Result is < 40.		
11/21/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g Note: <	5.06E+03	4.00E+01
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3518

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0761	0.0761
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	252.8	15.4
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code	Known µg of C from Original Sample	5.06E+3	<4.00E+1
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002589			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date	NOTE: FOR TOC: The Reported Result is	<	40.
11/21/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g Note:	<	3.20E+03 4.00E+01
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

NA

Date:

SAMPLE.WB1 REV 1.0

342100ML

3519

WORKBOOK PAGE: TRIPL5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0727	0.0727
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	282	37.3
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code	Known µg of C from Original Sample	5.06E+3	<4.00E+1
@TICTOC1	Duplicate Result	3.20E+3	<4.00E+1
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002589			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date	NOTE: FOR TOC: The Reported Result is	<	40.
11/21/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g Note:	<	3.75E+03 0.00E+00
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3520

WORKBOOK PAGE: SPIKE6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0580	0.0580
Work List	Final Coulometer Reading in µg (C1)	302.5	32.8
26749	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0737	0.0737
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	685.8	321.9
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/ml (SPK CONC)	602	3000
98005350	µg C in baseline (BL)	9.3	37.3
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002589			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time		TIC	TOC
05:30 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.08E+02	2.90E+03
BY-105	Percent Spike Recovery	101.0	96.8

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3521

WORKBOOK PAGE: SAM7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0463	0.0463
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	705	40
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002590			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g	1.50E+04	5.83E+01
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3522

WORKBOOK PAGE: DUP8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Size in g (SS)	TIC	TOC
DUP	Dilution Factor (DF)	0.0459	0.0459
Work List	μg of Carbon in Sample (C1)	603.3	41.7
26749	μg of Carbon from Baseline (C2)	9.3	37.3
Test Code	Known μg of C from Original Sample	1.50E+4	5.83E+1
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	μg of Carbon/g = (C1-C2) * DF / SS		
0	μg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	μg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002590			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:30 PM		5	40
Sample Point	μg of Carbon/g	1.29E+04	9.59E+01
BY-105			

Data Entered By:

CJO

Date: 11/25/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3523

WORKBOOK PAGE: SAM9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0278	0.0278
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	1115.8	141.8
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002591			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:30 PM		5	40
Sample Point	µg of Carbon/g	3.98E+04	3.76E+03
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3524

WORKBOOK PAGE: DUP10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0291	0.0291
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	1286.7	150.7
26749	µg of Carbon from Baseline (C2)	9.3	37.3
Test Code	Known µg of C from Original Sample	3.98E+4	3.76E+3
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98005350			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002591			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g	4.39E+04	3.90E+03
BY-105			

Data Entered By:

CJO

Date: 11/25/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3525

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 11/21/98

Time: 10:52:20

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis	Time	==== Coulometer	==== % Difference ==
1	0.51		0.40	0.00
2	1.01		0.90	55.56
3	1.50		15.30	94.12
4	2.00		101.20	84.88
5	2.50		250.30	59.57
6	3.00		404.50	38.12
7	3.50		497.00	18.61
8	4.00		540.90	8.12
9	4.50		562.40	3.82
10	5.00		572.30	1.73
11	5.50		577.30	0.87
12	6.00		580.30	0.52
13	6.50		581.80	0.26
14	7.00		583.10	0.22
15	7.50		584.10	0.17
16	8.00		584.90	0.14
17	8.50		585.60	0.12
18	9.00		586.20	0.10
19	9.50		586.90	0.12
20	10.00		587.40	0.09
21	10.50		588.00	0.10
22	11.00		588.60	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(588.6 - 37.27986) (1)/(200) = +2.757E+00 g/L Carbon

(588.6 - 37.27986) (1)/(200) (12) = +2.297E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 VERIFIED THE CALIBRATION ANALYSIS ON PAGES 3526 TO 3547.

Sample Run By:

SL HOOD

00002

200 uL = 26 N12-F

3526

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TIC/IOC REV 0.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/21/98

Time: 10:10:25

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.30	46.15
3	1.51	1.70	23.53
4	2.01	2.20	22.73
5	2.51	2.80	21.43
6	3.01	3.40	17.65
7	3.50	3.70	8.11
8	4.00	4.20	11.90
9	4.50	4.50	6.67
10	5.00	4.90	8.16
11	5.50	5.30	7.55
12	6.00	5.70	7.02
13	6.50	6.00	5.00
14	7.00	6.60	9.09
15	7.50	6.80	2.94
16	8.00	7.10	4.23
17	8.50	7.50	5.33
18	9.00	8.10	7.41
19	9.50	8.30	2.41
20	10.00	8.60	3.49
21	10.50	8.90	3.37
22	11.00	9.30	4.30

BLANK VALUE = 9.3 micrograms carbon

BLANK FACTOR = 9.3 / 10.99805 = +8.5E-01 ug/min Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.6

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/21/98

Time: 10:25:34

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.00	0.70	57.14
3	1.50	1.80	61.11
4	2.00	7.10	74.65
5	2.50	14.50	51.03
6	3.00	20.70	29.95
7	3.50	24.90	16.87
8	4.00	27.70	10.11
9	4.50	29.30	5.46
10	5.00	30.50	3.93
11	5.50	31.30	2.56
12	6.00	32.10	2.49
13	6.50	32.80	2.13
14	7.00	33.70	2.67
15	7.50	34.00	0.88
16	8.00	34.50	1.45
17	8.50	35.00	1.43
18	9.00	35.50	1.41
19	9.50	35.90	1.11
20	10.00	36.50	1.64
21	10.50	37.00	1.35
22	11.00	37.30	0.80

BLANK VALUE = 37.3 micrograms carbon

BLANK FACTOR = 37.3 / 10.99786 = +3.39E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 11/21/98

Time: 10:39:28

Sample Size = 1000 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .85 ug/minute C

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.50	0.00
2	1.00	1.00	50.00
3	1.50	7.90	87.34
4	2.00	108.10	92.69
5	2.50	262.20	58.77
6	3.00	398.00	34.12
7	3.50	493.00	19.27
8	4.00	547.50	9.95
9	4.50	576.20	4.98
10	5.00	590.50	2.42
11	5.50	597.70	1.20
12	6.00	601.20	0.58
13	6.50	603.10	0.32
14	7.00	604.30	0.20
15	7.50	605.10	0.13
16	8.00	605.90	0.13
17	8.50	606.60	0.12
18	9.00	607.00	0.07
19	9.50	607.80	0.13
20	10.00	608.10	0.05
21	10.50	608.60	0.08
22	11.00	609.10	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(609.1 - 9.348185) (1)/(1000) = +5.998E-01 g/L Carbon
 (609.1 - 9.348185) (1)/(1000) (12) = +4.998E-02 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = Inl = 25N/2-D

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.30

Sample: BLK 2

Date: 11/21/98

Time: 11:44:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.00	1.10	72.73
3	1.50	2.50	56.00
4	2.00	3.60	30.56
5	2.50	4.30	16.28
6	3.00	4.70	8.51
7	3.50	5.10	7.84
8	4.00	5.60	8.93
9	4.50	5.90	5.08
10	5.00	6.50	9.23
11	5.50	6.80	4.41
12	6.00	7.10	4.23
13	6.50	7.50	5.33
14	7.00	7.80	3.85
15	7.50	8.20	4.88
16	8.00	8.60	4.65
17	8.50	9.10	5.49
18	9.00	9.30	2.15
19	9.50	9.60	3.13
20	10.00	10.00	4.00
21	10.50	10.30	2.91
22	11.00	10.70	3.74

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(10.7 - 9.346628) (1)/(1) = +1.35E+00 g/L Carbon
 (10.7 - 9.346628) (1)/(1) (12) = +1.13E-01 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 11/21/98

Time: 11:56:56

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 3.39 ug/minute C

% Difference = 10

== Reading ==== Analysis Time ==== Coulcmeter ==== % Difference ==

1	0.51	0.40	0.00
2	1.00	0.70	42.86
3	1.50	2.40	70.83
4	2.00	6.90	65.22
5	2.50	11.50	40.00
6	3.00	15.50	25.81
7	3.50	17.70	12.43
8	4.00	19.20	7.81
9	4.50	20.40	5.88
10	5.00	21.30	4.23
11	5.50	21.90	2.74
12	6.00	22.50	2.67
13	6.50	23.00	2.17
14	7.00	23.40	1.71
15	7.50	23.90	2.09
16	8.00	24.30	1.65
17	8.50	24.80	2.02
18	9.00	25.10	1.20
19	9.50	25.50	1.57
20	10.00	25.90	1.54
21	10.50	26.20	1.15
22	11.00	26.60	1.50

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.98786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(26.6 - 37.27655) (1)/(1) =

< 5.00 E-3 g/L Carbon

(26.6 - 37.27655) (1)/(1)(12) =

< 4.17 E-4 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3531

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2-0

Sample: S98T002589

Date: 11/21/98

Time: 12:14:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	79.40	0.00
2	1.01	125.30	36.63
3	1.50	153.00	18.10
4	2.00	202.60	24.48
5	2.50	241.50	16.11
6	3.00	264.20	8.59
7	3.50	278.20	5.03
8	4.00	287.00	3.07
9	4.50	291.80	1.64
10	5.00	294.50	0.92
11	5.50	296.20	0.57
12	6.00	297.10	0.30
13	6.50	297.90	0.27
14	7.00	298.50	0.20
15	7.50	299.20	0.23
16	8.00	299.70	0.17
17	8.50	300.20	0.17
18	9.00	300.60	0.13
19	9.50	301.10	0.17
20	10.00	301.60	0.17
21	10.50	302.10	0.17
22	11.00	302.50	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(302.5 - 9.349014) (1) / (1)

+2.932E+02 g/L Carbon

(302.5 - 9.349014) (1) / (1) (12) =

+2.443E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .05809

3532

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002589

Date: 11/21/98

Time: 12:26:23

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.51	3.80	76.32
4	2.01	8.90	57.30
5	2.51	14.30	37.76
6	3.01	18.40	22.28
7	3.50	21.10	12.80
8	4.00	23.00	8.26
9	4.50	24.20	4.96
10	5.00	25.30	4.35
11	5.50	26.20	3.44
12	6.00	27.00	2.96
13	6.50	27.80	2.88
14	7.00	28.60	2.80
15	7.50	29.00	1.38
16	8.00	29.70	2.36
17	8.50	30.20	1.66
18	9.00	30.70	1.63
19	9.50	31.30	1.92
20	10.00	31.80	1.57
21	10.50	32.20	1.24
22	11.00	32.80	1.83

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(32.8 - 37.28959) (1)/(1) = < 5.00 E-3 g/L Carbon
 (32.8 - 37.28959) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

55- .0580g

3533

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC:REV 2.0

Sample: S98T02589 DUP

Date: 11/21/98

Time: 12:38:49

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C.

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis	Time	==== Coulometer	==== % Difference ==
1	0.51		0.40	0.00
2	1.00		0.89	50.00
3	1.50		12.50	93.60
4	2.00		103.70	87.95
5	2.50		165.10	37.19
6	3.00		200.30	17.57
7	3.50		221.30	9.49
8	4.00		233.20	5.10
9	4.50		239.30	2.55
10	5.00		243.70	1.81
11	5.50		245.60	0.77
12	6.00		247.40	0.73
13	6.50		248.40	0.40
14	7.00		249.10	0.28
15	7.50		249.80	0.28
16	8.00		250.20	0.16
17	8.50		250.90	0.28
18	9.00		251.10	0.08
19	9.50		251.60	0.20
20	10.00		252.10	0.20
21	10.50		252.50	0.16
22	11.00		252.80	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(252.8 - 9.349067) (1) / (1) = +2.435E+02 g/L Carbon
 (252.8 - 9.349067) (1) / (1) (12) = +2.029E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02589 DUP

Date: 11/21/98

Time: 14:20:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.50	0.20	0.00
2	1.00	0.50	60.00
3	1.50	0.70	28.57
4	2.00	1.00	30.00
5	2.50	1.30	23.08
6	3.00	1.50	13.33
7	3.50	1.80	16.67
8	4.00	2.40	25.00
9	4.50	3.50	31.43
10	5.00	4.90	28.57
11	5.50	6.30	22.22
12	6.00	8.20	23.17
13	6.50	9.30	11.83
14	7.00	10.70	13.08
15	7.50	11.50	6.96
16	8.00	12.20	5.74
17	8.50	12.90	5.43
18	9.00	13.50	4.44
19	9.50	14.10	4.26
20	10.00	14.60	3.42
21	10.50	15.10	3.31
22	11.00	15.40	1.95

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.94786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(15.4 - 37.28897) (1) / (1) =

< 5.00 E-3 g/L Carbon

(15.4 - 37.28897) (1) / (1) (12) =

< 4.17 E-4 Molar Carbon

<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

SL HOOD

00002

SS = .0761g

3535

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02589 TRP

Date: 11/21/98

Time: 14:50:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	1.90	0.00
2	1.01	3.10	38.71
3	1.50	22.00	85.91
4	2.00	126.30	82.58
5	2.50	188.80	33.10
6	3.00	225.90	16.42
7	3.50	248.00	8.91
8	4.00	261.50	5.16
9	4.50	265.20	2.86
10	5.00	273.40	1.54
11	5.50	275.80	0.87
12	6.00	277.00	0.43
13	6.50	277.70	0.25
14	7.00	278.70	0.36
15	7.50	278.90	0.07
16	8.00	279.40	0.18
17	8.50	279.80	0.14
18	9.00	280.30	0.18
19	9.50	280.70	0.14
20	10.00	281.20	0.18
21	10.50	281.50	0.11
22	11.00	282.00	0.18

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.98786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(282 - 9.348132) (1)/(1) =

+2.727E+02 g/L Carbon

(282 - 9.348132) (1)/(1) (12) =

+2.272E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0727g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV 2.0

Sample: S98T02589 TRP

Date: 11/21/98

Time: 15:03:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.80	37.50
3	1.50	3.00	73.33
4	2.00	9.60	68.75
5	2.50	16.00	40.00
6	3.00	20.90	23.44
7	3.50	24.40	14.34
8	4.00	26.30	7.22
9	4.50	28.00	6.07
10	5.00	29.40	4.76
11	5.50	30.20	2.65
12	6.00	31.10	2.89
13	6.50	32.00	2.81
14	7.00	32.80	2.44
15	7.50	33.50	2.09
16	8.00	34.20	2.05
17	8.50	34.70	1.44
18	9.00	35.40	1.98
19	9.50	35.80	1.12
20	10.00	36.30	1.38
21	10.50	36.70	1.09
22	11.00	37.30	1.61

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(37.3 - 37.27697) (1) / (1) =

+2.30E-02 g/L Carbon

(37.3 - 37.27697) (1) / (1) (12) =

+1.92E-03 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

07273

3537

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02589 SPK

Date: 11/21/98

Time: 15:16:16

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.00	0.80	50.00
3	1.50	16.90	95.27
4	2.00	149.70	88.71
5	2.50	323.70	53.75
6	3.00	462.90	30.07
7	3.50	555.30	16.64
8	4.00	612.80	9.38
9	4.50	646.20	5.17
10	5.00	664.40	2.74
11	5.50	673.20	1.31
12	6.00	677.50	0.63
13	6.50	679.90	0.35
14	7.00	680.90	0.15
15	7.50	682.00	0.16
16	8.00	682.60	0.09
17	8.50	683.40	0.12
18	9.00	683.90	0.07
19	9.50	684.30	0.06
20	10.00	684.80	0.07
21	10.50	685.30	0.07
22	11.00	685.80	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(685.8 - 9.346732) (1) (1)

(685.8 - 9.346732) (1) (1) (12)

+6.765E+02 g/L Carbon

+5.637E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0737g + 500ml spk 25N12-D

3538

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02589 ^{SPK} ~~SPK~~ _{SUB} Date: 11/21/98 Time: 15:29:21

Sample Size = 1 uL ¹¹⁻²¹⁻⁹⁸
Dil Factor = 1
Blank ID # =
Blank Value = 3.39 ug/minute C Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis	Time ==	Coulometer ==	% Difference ==
1	0.51		0.70	0.00
2	1.01		1.10	36.36
3	1.51		5.20	78.85
4	2.00		38.70	86.56
5	2.50		115.50	66.49
6	3.00		200.40	42.37
7	3.50		256.70	21.93
8	4.00		285.30	10.02
9	4.50		299.90	4.87
10	5.00		306.50	2.15
11	5.50		310.40	1.26
12	6.00		312.80	0.77
13	6.50		314.30	0.48
14	7.00		315.60	0.41
15	7.50		316.80	0.38
16	8.00		317.70	0.28
17	8.50		318.40	0.22
18	9.00		319.30	0.28
19	9.50		320.00	0.22
20	10.00		320.80	0.25
21	10.50		321.20	0.12
22	11.00		321.90	0.22

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(321.9 - 37.27986) (1) / (1) = +2.846E+02 g/L Carbon
 (321.9 - 37.27986) (1) / (1) (12) = +2.372E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0737g + .100 ml spk 26N12-F

3539

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002590-

Date: 11/21/98

Time: 15:50:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	51.60	0.00
2	1.00	78.10	33.93
3	1.50	102.40	23.73
4	2.00	276.00	62.90
5	2.50	453.30	39.11
6	3.00	555.90	18.46
7	3.50	619.00	10.19
8	4.00	656.90	5.77
9	4.50	678.90	3.24
10	5.00	689.20	1.49
11	5.50	694.00	0.69
12	6.00	696.60	0.37
13	6.50	698.30	0.24
14	7.00	699.50	0.17
15	7.50	700.50	0.14
16	8.00	701.30	0.11
17	8.50	702.00	0.10
18	9.00	702.70	0.10
19	9.50	703.30	0.09
20	10.00	703.80	0.07
21	10.50	704.50	0.10
22	11.00	705.00	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 =

+8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(705 - 9.347458) (1)/(1) =

+6.957E+02 g/L Carbon

(705 - 9.347458) (1)/(1) (12) =

+5.797E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = 0463g

3540

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: S98T002590

Date: 11/21/98

Time: 16:04:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.50	2.40	62.50
4	2.00	7.90	69.62
5	2.50	15.20	48.03
6	3.00	21.60	29.63
7	3.50	26.30	17.87
8	4.00	28.90	9.00
9	4.50	31.00	6.77
10	5.00	32.20	3.73
11	5.50	33.40	3.59
12	6.00	34.30	2.62
13	6.50	35.20	2.56
14	7.00	36.10	2.49
15	7.50	36.50	1.10
16	8.00	37.10	1.62
17	8.50	37.80	1.85
18	9.00	38.30	1.31
19	9.50	38.70	1.03
20	10.00	39.20	1.28
21	10.50	39.60	1.01
22	11.00	40.00	1.00

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(40 - 37.27986) (1)/(1) =

+2.72E+00 g/L Carbon

(40 - 37.27986) (1)/(1) (12) =

+2.27E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02590 DUP

Date: 11/21/98

Time: 16:16:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference ==
1	0.51		0.60	0.00
2	1.01		1.10	45.45
3	1.51		26.50	95.85
4	2.00		237.00	88.82
5	2.50		392.70	39.65
6	3.00		477.40	17.74
7	3.50		529.70	9.87
8	4.00		561.60	5.68
9	4.50		579.90	3.16
10	5.00		588.70	1.49
11	5.50		593.50	0.81
12	6.00		595.80	0.39
13	6.50		597.40	0.27
14	7.00		598.50	0.18
15	7.50		599.50	0.17
16	8.00		600.20	0.12
17	8.50		600.90	0.12
18	9.00		601.50	0.10
19	9.50		601.80	0.05
20	10.00		602.40	0.10
21	10.50		602.80	0.07
22	11.00		603.30	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786

+8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(603.3 - 9.348289) (1) (1) =
 (603.3 - 9.348289) (1) (1) (12) =

+5.940E+02 g/L Carbon
 +4.950E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0456g

3542

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TIC/TOC REV 2.0

Sample: S98T02590 DUP

Date: 11/21/98

Time: 16:30:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.01	0.00	0.00
2	0.51	0.50	100.00
3	1.00	0.90	44.44
4	1.50	3.70	75.68
5	2.00	10.90	66.06
6	2.50	18.40	40.76
7	3.00	24.40	24.59
8	3.50	28.80	15.28
9	4.00	31.50	8.57
10	4.50	33.50	5.97
11	5.00	35.00	4.29
12	5.50	35.90	2.51
13	6.00	36.80	2.45
14	6.50	37.80	2.65
15	7.00	38.30	1.31
16	7.50	38.80	1.29
17	8.00	39.40	1.52
18	8.50	39.90	1.25
19	9.00	40.50	1.48
20	9.50	40.90	0.98
21	10.00	41.40	1.21
22	10.50	41.70	0.72

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(41.7 - 35.58817) (1) / (1) = +6.11E+00 g/L Carbon
 (41.7 - 35.58817) (1) / (12) = +5.09E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

85 = .0456g

3543

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002591

Date: 11/21/98

Time: 16:48:27

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	85.90	0.00
2	1.01	124.80	31.17
3	1.51	196.00	36.33
4	2.00	543.50	63.94
5	2.50	783.50	30.63
6	3.00	920.70	14.90
7	3.50	1001.70	8.09
8	4.00	1029.80	4.58
9	4.50	1077.60	2.58
10	5.00	1091.30	1.26
11	5.50	1097.50	0.56
12	6.00	1101.20	0.34
13	6.50	1104.00	0.25
14	7.00	1107.00	0.27
15	7.50	1109.20	0.20
16	8.00	1110.60	0.13
17	8.50	1111.70	0.10
18	9.00	1112.60	0.08
19	9.50	1113.50	0.08
20	10.00	1114.30	0.07
21	10.50	1115.00	0.06
22	11.00	1115.80	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786 = +8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(1115.8 - 9.349689) (1)/(1) =

+1.1065E+03 g/L Carbon

(1115.8 - 9.349689) (1)/(1) (12) =

+9.2204E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

35 = .02783

3544

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002591

Date: 11/21/98

Time: 17:01:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	% Difference
1	0.51		0.90	0.00
2	1.01		1.40	35.71
3	1.50		16.20	91.36
4	2.00		58.10	72.12
5	2.50		90.40	35.73
6	3.00		111.00	18.56
7	3.50		122.20	9.17
8	4.00		127.70	4.31
9	4.50		130.80	2.37
10	5.00		132.70	1.43
11	5.50		134.20	1.12
12	6.00		135.30	0.81
13	6.50		136.10	0.59
14	7.00		137.00	0.66
15	7.50		137.70	0.51
16	8.00		138.60	0.65
17	8.50		139.20	0.43
18	9.00		139.80	0.43
19	9.50		140.30	0.36
20	10.00		140.80	0.36
21	10.50		141.40	0.42
22	11.00		141.80	0.28

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(141.8 - 37.28317) (1) / (1) = +1.045E+02 g/L Carbon
 (141.8 - 37.28317) (1) / (1) (12) = +8.710E+00 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0278g

3545

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02591 DUP

Date: 11/21/98

Time: 17:13:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.90	55.56
3	1.50	27.50	96.73
4	2.00	421.90	93.48
5	2.50	804.00	47.52
6	3.00	1002.20	19.78
7	3.50	1116.70	10.25
8	4.00	1184.10	5.69
9	4.50	1226.10	3.43
10	5.00	1251.70	2.05
11	5.50	1265.40	1.08
12	6.00	1272.90	0.59
13	6.50	1277.30	0.34
14	7.00	1279.60	0.18
15	7.50	1281.10	0.12
16	8.00	1282.30	0.09
17	8.50	1283.30	0.08
18	9.00	1284.10	0.06
19	9.50	1284.80	0.05
20	10.00	1285.50	0.05
21	10.50	1286.20	0.05
22	11.00	1286.70	0.04

USER INPUT BLANK VALUE

BLANK VALUE = 9.348185 micrograms carbon

BLANK FACTOR = 9.348185 / 10.99786

+8.5E-01 ug/min Carbon

SAMPLE RESULTS:

(1286.7 - 9.348185) (1) / (1) =

(1286.7 - 9.348185) (1) / (1) (12) =

+1.2774E+03 g/L Carbon

+1.0645E+02 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02591 DUP

Date: 11/21/98

Time: 17:26:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.39 ug/minute.C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.70	0.00
2	1.01	1.20	41.67
3	1.51	5.30	77.36
4	2.00	53.60	90.11
5	2.50	96.40	44.40
6	3.00	113.60	15.14
7	3.50	121.20	6.27
8	4.00	125.90	3.73
9	4.50	129.20	2.55
10	5.00	132.70	2.64
11	5.50	136.00	2.43
12	6.00	139.20	2.30
13	6.50	141.50	1.63
14	7.00	143.20	1.19
15	7.50	144.70	1.04
16	8.00	146.00	0.89
17	8.50	147.00	0.68
18	9.00	147.80	0.54
19	9.50	148.80	0.67
20	10.00	149.40	0.40
21	10.50	150.00	0.40
22	11.00	150.70	0.46

USER INPUT BLANK VALUE

BLANK VALUE = 37.28276 micrograms carbon

BLANK FACTOR = 37.28276 / 10.99786 = +3.4E+00 ug/min Carbon

SAMPLE RESULTS:

(150.7 - 37.28917) (1)/(1) = +1.134E+02 g/L Carbon
 (150.7 - 37.28917) (1)/(1) (12) = +9.451E+00 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = 0.02913

LABCORE Completed Worklist Report for Worklist# 26750

Analyst: slh

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	6.02E+02	5.93E+2	98.505 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.78E+3	92.667 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	2.70E+0	2.700 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	2.70E+0	2.700 ug/g	
3 SAMPLE	S98T002592	0	@TICTOC1 TIC-02	SOLID	N/A	5.33E+03	5.000 ug/g	
3 SAMPLE	S98T002592	0	@TICTOC1 TOC-02	SOLID	N/A	9.90E+2	40.000 ug/g	
4 DUP	S98T002592	0	@TICTOC1 TIC-02	SOLID	5.33E+3	5.82E+3	8.789 RPD	
4 DUP	S98T002592	0	@TICTOC1 TOC-02	SOLID	9.90E+2	7.67E+2	25.384 RPD	
5 TRIPL	S98T002592	0	@TICTOC1 TIC-02	SOLID	5.33E+3	3.47E+3	42.273 RPD	
5 TRIPL	S98T002592	0	@TICTOC1 TOC-02	SOLID	9.90E+2	1.01E+3	2.000 RPD	
6 SAMPLE	S98T002593	0	@TICTOC1 TIC-02	SOLID	N/A	3.91E+03	5.000 ug/g	
6 SAMPLE	S98T002593	0	@TICTOC1 TOC-02	SOLID	N/A	1.88E+04	40.000 ug/g	
7 DUP	S98T002593	0	@TICTOC1 TIC-02	SOLID	3.91E+3	3.95E+3	1.018 RPD	
7 DUP	S98T002593	0	@TICTOC1 TOC-02	SOLID	1.88E+4	1.63E+4	14.245 RPD	
8 SPK	S98T002593	0	@TICTOC1 TIC-02	SOLID	1.00E+02	1.08E+2	108.000 % Recovery	
8 SPK	S98T002593	0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.01E+2	101.000 % Recovery	
9 SAMPLE	S98T002594	0	@TICTOC1 TIC-02	SOLID	N/A	5.04E+03	5.000 ug/g	
9 SAMPLE	S98T002594	0	@TICTOC1 TOC-02	SOLID	N/A	2.16E+04	40.000 ug/g	
10 DUP	S98T002594	0	@TICTOC1 TIC-02	SOLID	5.04E+3	5.19E+3	2.933 RPD	
10 DUP	S98T002594	0	@TICTOC1 TOC-02	SOLID	2.16E+4	2.08E+4	3.774 RPD	

Final page for worklist# 26750

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

12/1/98
Date

LABCORE Completed Worklist Report for Worklist# 26750

Analyst: slh Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	6.02E+02	5.93E+2	98.505 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.78E+3	92.667 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	2.70E+0	2.700 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	2.61E+0	2.610 ug/g	
3 SAMPLE	S98T002592	0	@TICTOC1 TIC-02	SOLID	N/A	5.33E+03	5.000 ug/g	
3 SAMPLE	S98T002592	0	@TICTOC1 TOC-02	SOLID	N/A	9.88E+02	40.000 ug/g	
4 DUP	S98T002592	0	@TICTOC1 TIC-02	SOLID	5.33E+3	5.82E+3	8.789 RPD	
4 DUP	S98T002592	0	@TICTOC1 TOC-02	SOLID	9.88E+2	7.65E+2	25.442 RPD	
5 TRIPL	S98T002592	0	@TICTOC1 TIC-02	SOLID	5.33E+3	3.47E+3	42.273 RPD	
5 TRIPL	S98T002592	0	@TICTOC1 TOC-02	SOLID	9.88E+2	1.01E+3	2.202 RPD	
6 SAMPLE	S98T002593	0	@TICTOC1 TIC-02	SOLID	N/A	3.91E+03	5.000 ug/g	
6 SAMPLE	S98T002593	0	@TICTOC1 TOC-02	SOLID	N/A	1.88E+04	40.000 ug/g	
7 DUP	S98T002593	0	@TICTOC1 TIC-02	SOLID	3.91E+3	3.95E+3	1.018 RPD	
7 DUP	S98T002593	0	@TICTOC1 TOC-02	SOLID	1.88E+4	1.63E+4	14.245 RPD	
8 SPK	S98T002593	0	@TICTOC1 TIC-02	SOLID	1.00E+02	1.08E+02	108.000 % Recovery	
8 SPK	S98T002593	0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.01E+02	101.000 % Recovery	
9 SAMPLE	S98T002594	0	@TICTOC1 TIC-02	SOLID	N/A	5.04E+03	5.000 ug/g	
9 SAMPLE	S98T002594	0	@TICTOC1 TOC-02	SOLID	N/A	2.16E+04	40.000 ug/g	
10 DUP	S98T002594	0	@TICTOC1 TIC-02	SOLID	5.04E+3	5.19E+3	2.933 RPD	
10 DUP	S98T002594	0	@TICTOC1 TOC-02	SOLID	2.16E+4	2.08E+4	3.774 RPD	

Final page for worklist# 26750

Analyst Signature _____ Date _____

C. J. O'Neil 11/25/98
Analyst Signature _____ Date _____

[Signature] 12/1/98
Reviewer Signature _____ Date _____

10/23/98 09:37
A-0604-1

Page: 1

LABCORE Data Entry Template for Worklist# 26750

Analyst: SLH Instrument: CARB2 _____ Book# TLC 25N/2-D
 Method: LA-342-100 Rev/Mod F-1 TOC 26N/2-F
 Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	SOLID		
2 BLNK			@TICTOC1	SOLID		
3 SAMPLE	S98T002592 0		@TICTOC1	SOLID	98000436 BY-105 (2)	
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002592 0		@TICTOC1	SOLID		
5 SAMPLE	S98T002593 0		@TICTOC1	SOLID	98000436 BY-105 (2)	
	Analytes Requested: TIC-02 , TOC-02					
6 DUP	S98T002593 0		@TICTOC1	SOLID		
7 SPK	S98T002593 0		@TICTOC1	SOLID		
8 SAMPLE	S98T002594 0		@TICTOC1	SOLID	98000436 BY-105 (2)	
	Analytes Requested: TIC-02 , TOC-02					
9 DUP	S98T002594 0		@TICTOC1	SOLID		

Final page for worklist # 26750

Sandra H. H. B. B. B.
 Analyst Signature Date
 11-22-98

C. J. P. P. P.
 Analyst Signature Date
 11/25/98

Fixed
 out
 11-22-98
 SLH

Data Entry Comments:
 Patch 5351

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	11.9	27.3
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code			
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	5	40
04:00 PM			
Sample Point	µg of Carbon	2.70E+00	2.70E+00
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>N/A</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3551

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	602.4	581.5
26750	μg of Carbon from Baseline (C2)	9.2	24.6
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98005351			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
04:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.93E+02	2.78E+03
BY-105	Percent Standard Recovery	98.5	92.8

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	N/A	Date:	

STANDARD.WB1 REV 1.0

342100ML

3552

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0399	0.0399
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	222	64.1
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002592			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	5.33E+03	9.90E+02
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0417	0.0417
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	252	56.6
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code	Known µg of C from Original Sample	5.33E+3	9.88E+2
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002592			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	5.82E+03	7.67E+02
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

WORKBOOK PAGE: TRIPL5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0342	0.0342
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	127.9	59.2
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code	Known µg of C from Original Sample	5.33E+3	9.88E+2
@TIC/TOC1	Duplicate Result	5.82E+3	7.65E+2
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002592			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	3.47E+03	1.01E+03
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3555

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0428	0.0428
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	176.6	830.4
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002593			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	3.91E+03	1.88E+04
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0415	0.0415
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	173	700.2
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code	Known µg of C from Original Sample	3.91E+3	1.88E+4
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002593			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	3.95E+03	1.63E+04
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3557

WORKBOOK PAGE: SPIKE8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0428	0.0428
Work List	Final Coulometer Reading in μg (C1)	176.6	830.4
26750	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0385	0.0385
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in μg (C2)	485.8	1051.2
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in $\mu\text{g/mL}$ (SPK CONC)	602	3000
98005351	$\mu\text{g C}$ in baseline (BL)	9.2	24.6
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = $((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100$		
S98T002593			
Instrument Code	QC Actual in $\mu\text{g/mL}$ = Spike Value ($\mu\text{g/mL}$)		
CARB2	QC Found in $\mu\text{g/mL}$ = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time		TIC	TOC
04:00 PM	QC Actual in $\mu\text{g/mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g/mL}$	6.52E+02	3.02E+03
BY-105	Percent Spike Recovery	108.3	100.6

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SPIKE.WB1 REV 1.3

342100ML

3558

WORKBOOK PAGE: SAM9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0418	0.0418
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	219.7	929.1
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002594			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	5.04E+03	2.16E+04
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3559

WORKBOOK PAGE: DUP10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0506	0.0506
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	271.9	1077.9
26750	µg of Carbon from Baseline (C2)	9.2	24.6
Test Code	Known µg of C from Original Sample	5.04E+3	2.16E+4
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005351			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002594			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
11/25/98			
Analysis Date			
11/22/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	5.19E+03	2.08E+04
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3560

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV 2.0

Sample: STD 2

Date: 11/22/98

Time: 09:19:59

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst

SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	1.60	0.00
2	1.00	10.50	84.76
3	1.50	105.10	90.01
4	2.00	257.50	59.18
5	2.50	393.20	34.51
6	3.00	484.80	18.89
7	3.50	540.50	10.31
8	4.00	540.30	5.23
9	4.50	584.60	2.45
10	5.00	591.40	1.15
11	5.50	594.20	0.47
12	6.00	595.90	0.29
13	6.50	597.10	0.20
14	7.00	598.00	0.15
15	7.50	598.70	0.12
16	8.00	599.20	0.08
17	8.50	599.80	0.10
18	9.00	600.30	0.08
19	9.50	601.00	0.12
20	10.00	601.50	0.08
21	10.50	601.90	0.07
22	11.00	602.40	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 =

+8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(602.4 - 9.240563) (1) / (1000) =

+5.932E-01 g/L Carbon

(602.4 - 9.240563) (1) / (1000) (12) =

+4.943E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 356 TO 358.

Sample Run By: *SL Hood*

SL HOOD

00002

SS-1ml - 25N12-D

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC RATIO 2.0

Sample: STD 2

Date: 11/22/98

Time: 09:37:47

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Concentration	% Difference
1	0.51	0.40	0.00
2	1.01	1.50	73.33
3	1.50	14.50	89.66
4	2.00	77.20	79.63
5	2.50	191.40	62.80
6	3.00	348.60	45.09
7	3.50	465.40	25.10
8	4.00	524.70	11.30
9	4.50	552.70	5.07
10	5.00	565.10	2.19
11	5.50	570.70	0.98
12	6.00	573.50	0.49
13	6.50	575.00	0.26
14	7.00	576.10	0.19
15	7.50	577.20	0.19
16	8.00	577.90	0.12
17	8.50	578.50	0.10
18	9.00	579.20	0.12
19	9.50	579.80	0.10
20	10.00	580.40	0.10
21	10.50	581.00	0.10
22	11.00	581.50	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms Carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(581.5 - 24.63522) (1)/(200) =

+2.784E+00 g/L Carbon

(581.5 - 24.63522) (1)/(200) (12) =

+2.320E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!!!>>>>

Sample Run By:

SL HOOD

00002

3562

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TIC/IOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/22/98

Time: 08:26:45

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	% Difference
1	0.51		2.60	0.00
2	1.01		2.60	7.14
3	1.51		3.10	9.68
4	2.01		3.60	13.89
5	2.51		4.00	10.00
6	3.01		4.50	11.11
7	3.51		4.90	8.16
8	4.00		5.10	3.92
9	4.50		5.40	5.56
10	5.00		5.70	5.26
11	5.50		6.00	5.00
12	6.00		6.30	4.76
13	6.50		6.60	4.55
14	7.00		6.80	2.94
15	7.50		7.10	4.23
16	8.00		7.40	4.05
17	8.50		7.70	3.90
18	9.00		8.00	3.75
19	9.50		8.30	3.61
20	10.00		8.50	2.35
21	10.50		8.80	3.41
22	11.00		9.20	4.35

BLANK VALUE = 9.2 micrograms carbon

BLANK FACTOR = 9.2 / 10.99969 =

+8.4E-01

ug/min Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TIC/TOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/22/98

Time: 08:39:06

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	Difference
1	0.51		0.30	0.00
2	1.01		0.80	62.50
3	1.50		2.60	77.78
4	2.00		7.50	52.63
5	2.50		11.20	32.14
6	3.00		13.50	17.04
7	3.50		15.30	11.76
8	4.00		16.50	7.27
9	4.50		17.50	5.71
10	5.00		18.60	5.91
11	5.50		19.50	4.62
12	6.00		20.10	2.99
13	6.50		20.70	2.90
14	7.00		21.30	2.82
15	7.50		21.76	1.84
16	8.00		22.20	2.25
17	8.50		22.60	1.77
18	9.00		23.10	2.16
19	9.50		23.60	2.12
20	10.00		23.90	1.26
21	10.50		24.80	1.65
22	11.00		24.60	1.22

BLANK VALUE = 24.6 micrograms carbon

BLANK FACTOR = $24.6 / 10.99969 = +2.24E+00$ ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: BLK 2

Date: 11/22/98

Time: 9:52:29

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	Difference ==
1	0.51	0.50	0.00
2	1.01	1.10	54.55
3	1.50	2.60	57.69
4	2.00	3.80	31.58
5	2.50	4.60	17.39
6	3.00	5.10	9.80
7	3.50	5.70	10.53
8	4.00	6.30	9.52
9	4.50	6.70	5.97
10	5.00	7.00	4.29
11	5.50	7.70	9.09
12	6.00	7.90	2.53
13	6.50	8.30	4.82
14	7.00	8.80	5.68
15	7.50	9.10	3.30
16	8.00	9.50	4.21
17	8.50	10.00	5.00
18	9.00	10.40	3.85
19	9.50	10.80	3.70
20	10.00	11.10	2.70
21	10.50	11.60	4.31
22	11.00	11.90	2.52

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988

+8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(11.9 - 9.239025) (1) / (1)

+2.66E+00 g/L Carbon

(11.9 - 9.239025) (1) / (1) (12)

+2.22E-01 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 11/22/98

Time: 10:05:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	Difference
1	0.51	0.40	0.00
2	1.01	3.70	89.19
3	1.50	9.00	58.89
4	2.00	13.00	30.77
5	2.50	15.80	17.72
6	3.00	17.90	11.73
7	3.50	19.20	6.77
8	4.00	20.10	4.48
9	4.50	20.80	3.37
10	5.00	21.40	2.80
11	5.50	21.90	2.28
12	6.00	22.50	2.67
13	6.50	23.00	2.17
14	7.00	23.70	2.95
15	7.50	24.00	1.25
16	8.00	24.70	2.83
17	8.50	24.90	0.80
18	9.00	25.40	1.97
19	9.50	25.90	1.93
20	10.00	26.30	1.52
21	10.50	26.70	1.50
22	11.00	27.30	2.20

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(27.3 - 24.63522) (1)/(1) = +2.66E+00 g/L Carbon
 (27.3 - 24.63522) (1)/(1) (1) = +2.22E-01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T0092592

Date: 11/22/98

Time: 10:18:53

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.50	0.00
2	1.01	1.40	64.29
3	1.50	29.70	95.29
4	2.00	100.90	70.56
5	2.50	148.10	31.87
6	3.01	178.30	16.84
7	3.50	196.70	9.46
8	4.00	207.20	5.07
9	4.50	212.10	2.31
10	5.00	214.80	1.26
11	5.50	216.20	0.65
12	6.00	217.20	0.46
13	6.50	217.90	0.32
14	7.00	218.50	0.27
15	7.50	218.90	0.18
16	8.00	219.40	0.23
17	8.50	219.80	0.18
18	9.00	220.40	0.27
19	9.50	220.80	0.18
20	10.00	221.20	0.18
21	10.50	221.60	0.18
22	11.00	222.00	0.18

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms Carbon

BLANK FACTOR = 9.239897 / 10.99988 = +8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(222 - 9.238975) (1)/(1) = +2.128E+02 g/L Carbon
 (222 - 9.238975) (1)/(1) (12) = +1.773E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

55 = .03998

3567

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV. 2.0

Sample: S98T002592

Date: 11/22/98

Time: 10:33:06

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst

SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== Difference ==
1	0.51	0.40	0.00
2	1.01	1.00	60.00
3	1.50	13.30	92.48
4	2.00	29.60	55.07
5	2.50	40.60	27.09
6	3.00	46.90	13.43
7	3.50	50.60	7.31
8	4.00	52.90	4.35
9	4.50	54.60	3.11
10	5.00	55.80	2.15
11	5.50	56.90	1.93
12	6.00	57.90	1.73
13	6.50	58.60	1.19
14	7.00	59.50	1.51
15	7.50	60.20	1.16
16	8.00	60.90	1.15
17	8.50	61.50	0.98
18	9.00	62.00	0.81
19	9.50	62.50	0.80
20	10.00	63.00	0.79
21	10.50	63.50	0.79
22	11.00	64.10	0.94

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(64.1 - 24.63522) (1)/(1) =

+3.95E+01

g/L Carbon

(64.1 - 24.63522) (1)/(1) (12) =

+3.29E+00

Molar Carbon

<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

SL HOOD

00002

SS = .0399g

3568

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02592 DUP

Date: 11/22/98

Time: 10:47:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst

SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Colloimeter	% Difference ==
1	0.51	0.30	0.00
2	1.01	0.80	62.50
3	1.50	21.80	96.33
4	2.00	92.80	76.51
5	2.50	156.10	40.55
6	3.00	196.20	20.44
7	3.50	221.10	11.26
8	4.00	234.40	5.67
9	4.50	240.90	2.70
10	5.00	244.30	1.39
11	5.50	246.00	0.69
12	6.00	247.00	0.40
13	6.50	247.80	0.32
14	7.00	248.60	0.32
15	7.50	249.00	0.16
16	8.00	249.40	0.16
17	8.50	249.90	0.20
18	9.00	250.40	0.20
19	9.50	250.80	0.16
20	10.00	251.20	0.16
21	10.50	251.60	0.16
22	11.00	252.00	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 = +8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(252 - 9.236821) (1)/(1) =

-2.428E+02 g/L Carbon

(252 - 9.236821) (1)/(1) (12) =

-2.023E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV 2.0

Sample: S98T02592 DUP

Date: 11/22/98

Time: 11:00:16

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.50	73.33
3	1.51	11.30	86.73
4	2.01	24.40	53.69
5	2.51	33.40	26.95
6	3.00	39.70	15.87
7	3.50	43.40	8.53
8	4.00	45.90	5.45
9	4.50	47.60	3.57
10	5.00	48.90	2.66
11	5.50	49.90	2.00
12	6.00	51.00	2.16
13	6.50	51.40	0.78
14	7.00	52.20	1.53
15	7.50	52.90	1.32
16	8.00	53.70	1.49
17	8.50	54.40	1.29
18	9.00	54.70	0.55
19	9.50	55.20	0.91
20	10.00	55.70	0.90
21	10.50	56.20	0.89
22	11.00	56.60	0.71

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(56.6 - 24.6374) (1)/(1) = +3.20E+01 g/L Carbon
 (56.6 - 24.6374) (1)/(1) (12) = -2.66E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0417g

3570

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02592 TRP

Date: 11/22/98

Time: 14:00:37

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Conometer ==	% Difference ==
1	0.51	0.30	0.00
2	1.00	1.00	70.00
3	1.50	10.90	90.83
4	2.00	45.00	76.25
5	2.50	90.70	43.12
6	3.00	102.80	21.50
7	3.50	113.30	9.27
8	4.00	118.10	4.06
9	4.50	120.90	2.32
10	5.00	122.00	0.90
11	5.50	123.10	0.89
12	6.00	123.60	0.40
13	6.50	124.10	0.40
14	7.00	124.70	0.48
15	7.50	125.00	0.24
16	8.00	125.50	0.40
17	8.50	125.90	0.32
18	9.00	126.30	0.32
19	9.50	126.70	0.32
20	10.00	127.10	0.31
21	10.50	127.40	0.24
22	11.00	127.90	0.39

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 = +8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(127.9 - 9.238205) (1) / (1) = +1.187E+02 g/L Carbon
 (127.9 - 9.238205) (1) / (1) (12) = +9.888E+00 Molar Carbon

Sample Run By:

SL HOOD

00002

SS-0342g

3571

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/IOC REV 2.0

Sample: S98T02592 TRP

Date: 11/22/98

Time: 14:13:07

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst

SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== Difference ==
1	0.51	0.40	0.00
2	1.01	1.40	71.43
3	1.50	14.10	90.07
4	2.00	27.40	48.54
5	2.50	36.60	25.14
6	3.00	42.80	13.48
7	3.50	45.90	7.84
8	4.00	48.50	5.36
9	4.50	49.80	2.61
10	5.00	51.10	2.54
11	5.50	52.20	2.11
12	6.00	53.20	1.88
13	6.50	54.10	1.66
14	7.00	55.00	1.64
15	7.50	55.50	0.90
16	8.00	56.30	1.42
17	8.50	56.70	0.71
18	9.00	57.30	1.05
19	9.50	57.80	0.87
20	10.00	58.20	0.69
21	10.50	58.70	0.85
22	11.00	59.20	0.84

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(59.2 - 24.63508) (1) / (1)

+3.46E+01

g/L Carbon

(59.2 - 24.63508) (1) / (1) (12)

+2.88E+00

Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0342g

3572

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: S98T002593

Date: 11/22/98

Time: 11:12:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst

SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.00	60.00
3	1.50	3.30	69.70
4	2.00	18.00	81.67
5	2.50	49.20	63.41
6	3.00	86.00	42.79
7	3.50	118.10	27.18
8	4.00	144.40	16.48
9	4.50	158.00	8.77
10	5.00	162.10	4.38
11	5.50	166.20	2.47
12	6.00	168.30	1.25
13	6.50	169.70	0.82
14	7.00	170.80	0.64
15	7.50	171.70	0.52
16	8.00	173.50	0.46
17	8.50	173.50	0.58
18	9.00	174.10	0.34
19	9.50	174.70	0.34
20	10.00	175.40	0.40
21	10.50	176.10	0.40
22	11.00	176.60	0.28

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 =

+8.4E-01

ug/min Carbon

SAMPLE RESULTS:

(176.6 - 9.239795) (1)/(1) =

+1.674E+02 g/L Carbon

(176.6 - 9.239795) (1)/(1) (12) =

+1.395E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0428g

3573

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV 2.0

Sample: S98T002593

Date: 11/22/98

Time: 13:37:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	Difference ==
1	0.51	0.70	0.00
2	1.00	2.30	69.57
3	1.50	30.50	92.46
4	2.00	37.50	68.72
5	2.50	202.00	51.73
6	3.00	317.30	35.93
7	3.50	415.50	24.12
8	4.00	492.10	15.57
9	4.50	554.00	11.17
10	5.00	605.80	8.55
11	5.50	646.60	6.31
12	6.00	681.10	5.07
13	6.50	710.30	4.11
14	7.00	731.60	3.31
15	7.50	754.10	2.71
16	8.00	771.60	2.14
17	8.50	785.90	1.82
18	9.00	797.60	1.47
19	9.50	807.50	1.23
20	10.00	816.20	1.07
21	10.50	823.40	0.87
22	11.00	830.40	0.84

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(830.4 - 24.6333) (1) / (1) = +8.058E+02 g/L Carbon
 (830.4 - 24.6333) (1) / (1) (12) = +6.715E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0428g

3574

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICIDC R/W 0.0

Sample: S98T02593 DUP

Date: 11/22/98

Time: 14:25:12

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	Difference
1	0.51		0.50	0.00
2	1.00		1.10	54.55
3	1.50		4.50	75.56
4	2.00		16.20	72.22
5	2.50		39.60	59.09
6	3.00		72.60	45.45
7	3.50		107.80	32.65
8	4.00		134.30	19.73
9	4.50		149.40	10.11
10	5.00		157.30	5.02
11	5.50		161.60	2.66
12	6.00		164.00	1.46
13	6.50		165.60	0.97
14	7.00		166.90	0.78
15	7.50		168.00	0.65
16	8.00		168.70	0.41
17	8.50		169.50	0.47
18	9.00		170.30	0.47
19	9.50		171.00	0.41
20	10.00		171.70	0.41
21	10.50		172.50	0.46
22	11.00		173.00	0.29

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 =

+8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(173 - 9.237488) (1)/(1)

(173 - 9.237488) (1)/(1) (12) =

+1.638E+02 g/L Carbon

+1.365E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0415g

3575

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02593 DUP

Date: 11/22/98

Time: 14:38:21

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute.C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference ==
1	0.51		0.70	0.00
2	1.01		2.40	70.83
3	1.50		31.00	92.26
4	2.00		87.00	64.37
5	2.50		168.20	48.28
6	3.00		259.40	35.16
7	3.50		342.20	24.20
8	4.00		409.90	16.52
9	4.50		462.60	11.39
10	5.00		506.90	8.74
11	5.50		542.50	6.56
12	6.00		572.00	5.16
13	6.50		596.60	4.12
14	7.00		616.80	3.27
15	7.50		634.00	2.71
16	8.00		648.10	2.18
17	8.50		660.60	1.89
18	9.00		670.60	1.49
19	9.50		679.80	1.35
20	10.00		687.90	1.18
21	10.50		694.50	0.95
22	11.00		700.20	0.81

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(700.2 - 24.63152) (1) / (1) = +6.756E+02 g/L Carbon
 (700.2 - 24.63152) (1) / (1) (12) = +5.630E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

86 - .0415g

3576

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TIC/OC REV 2.0

Sample: S98T02593 SPK

Date: 11/22/98

Time: 14:54:44

Sample Size = 1 uL

Dil Factor = 1

Blank ID #

Blank Value = .84 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	0.70	0.00
2	1.00	1.50	53.33
3	1.51	23.60	93.64
4	2.00	111.40	78.82
5	2.50	224.30	50.33
6	3.00	318.30	29.53
7	3.50	385.50	17.43
8	4.00	429.20	10.18
9	4.50	454.70	5.61
10	5.00	466.60	2.55
11	5.50	472.00	1.14
12	6.00	475.00	0.63
13	6.50	477.10	0.44
14	7.00	478.70	0.33
15	7.50	479.60	0.19
16	8.00	480.80	0.25
17	8.50	481.80	0.21
18	9.00	482.60	0.17
19	9.50	483.50	0.19
20	10.00	484.50	0.21
21	10.50	485.20	0.14
22	11.00	485.80	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 =

+8.4E-01

ug/min Carbon

SAMPLE RESULTS:

(485.8 - 9.237488) (1)/(1) =

+1.766E+02 g/L Carbon

(485.8 - 9.237488) (1)/(1) (12) =

+1.971E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SGC .0385g + .500L spk 25/11/98

3577

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02593 SPK

Date: 11/22/98

Time: 15:06:58

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	3.40	79.41
3	1.50	53.50	93.64
4	2.00	159.50	66.46
5	2.50	335.00	52.39
6	3.00	524.00	36.07
7	3.50	653.50	19.82
8	4.00	738.00	11.45
9	4.50	799.50	7.69
10	5.00	846.50	5.55
11	5.50	883.80	4.22
12	6.00	914.40	3.35
13	6.50	939.50	2.67
14	7.00	960.50	2.19
15	7.50	978.70	1.86
16	8.00	993.90	1.53
17	8.50	1007.20	1.32
18	9.00	1018.70	1.13
19	9.50	1028.50	0.95
20	10.00	1036.90	0.81
21	10.50	1044.60	0.74
22	11.00	1051.20	0.63

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(1051.2 - 24.6333) (1) / (1) = +1.0266E+03 g/L Carbon
 (1051.2 - 24.6333) (1) / (1) (12) = +8.5547E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0385g + .100 ml spk 26.12 F

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

HNH-1663 REV. 0

Sample: S98T002594

Date: 11/22/98

Time: 15:19:05

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference ==
1	0.51		5.40	0.00
2	1.01		9.90	45.45
3	1.50		15.40	35.71
4	2.00		34.60	55.49
5	2.50		67.00	48.36
6	3.00		108.60	38.31
7	3.50		149.60	27.41
8	4.00		179.70	16.75
9	4.50		196.50	8.55
10	5.00		203.70	3.53
11	5.50		207.40	1.78
12	6.00		209.80	1.14
13	6.50		211.30	0.71
14	7.00		212.60	0.61
15	7.50		213.70	0.51
16	8.00		214.60	0.42
17	8.50		215.70	0.51
18	9.00		216.40	0.32
19	9.50		217.40	0.46
20	10.00		218.10	0.32
21	10.50		219.00	0.41
22	11.00		219.70	0.32

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 =

+8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(219.7 - 9.239025) (1) / (1) =

+2.105E+02 g/L Carbon

(219.7 - 9.239025) (1) / (1) (12) =

+1.754E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0418g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TIC/TOC REV 3.0

Sample: S98T002594

Date: 11/22/98

Time: 15:31:18

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.00	4.10	85.37
3	1.50	73.90	94.45
4	2.00	191.00	61.31
5	2.50	326.00	41.41
6	3.00	453.00	28.04
7	3.50	553.00	18.08
8	4.00	628.70	12.04
9	4.50	686.60	8.43
10	5.00	733.10	6.34
11	5.50	771.40	4.97
12	6.00	802.70	3.90
13	6.50	828.10	3.07
14	7.00	848.60	2.42
15	7.50	865.40	1.94
16	8.00	879.60	1.61
17	8.50	891.60	1.35
18	9.00	902.00	1.15
19	9.50	910.70	0.96
20	10.00	917.70	0.76
21	10.50	923.80	0.66
22	11.00	929.10	0.57

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(929.1 - 24.63508) (1)/(1) = +9.045E+02 g/L Carbon
 (929.1 - 24.63508) (1)/(1) (12) = +7.537E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .04Bg

3580

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02594 DUP

Date: 11/22/98

Time: 15:43:21

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .84 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.50	3.80	0.00
2	1.00	6.30	39.68
3	1.50	10.90	42.20
4	2.00	30.40	64.14
5	2.50	66.70	54.42
6	3.00	108.60	38.58
7	3.50	155.60	30.21
8	4.00	201.50	22.78
9	4.50	232.20	13.22
10	5.00	248.90	6.71
11	5.50	256.40	2.93
12	6.00	260.10	1.42
13	6.50	262.30	0.84
14	7.00	264.10	0.68
15	7.50	265.30	0.45
16	8.00	266.30	0.38
17	8.50	267.40	0.41
18	9.00	268.40	0.37
19	9.50	269.40	0.37
20	10.00	270.20	0.30
21	10.50	271.10	0.33
22	11.00	271.90	0.29

USER INPUT BLANK VALUE

BLANK VALUE = 9.239897 micrograms carbon

BLANK FACTOR = 9.239897 / 10.99988 = +8.4E-01 ug/min Carbon

SAMPLE RESULTS:

(271.9 - 9.237488) (1)/(1) =

+2.627E+02 g/L Carbon

(271.9 - 9.237488) (1)/(1) (12) =

+2.189E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3581

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02594 DUP

Date: 11/22/98

Time: 15:56:01

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.24 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.80	0.00
2	1.01	2.70	70.37
3	1.50	44.10	93.88
4	2.00	135.90	67.55
5	2.50	272.60	50.15
6	3.00	426.00	36.01
7	3.50	559.50	23.86
8	4.00	664.30	15.78
9	4.50	745.80	10.93
10	5.00	810.00	7.93
11	5.50	863.60	6.21
12	6.00	906.90	4.77
13	6.50	941.90	3.72
14	7.00	970.90	2.99
15	7.50	994.00	2.32
16	8.00	1013.10	1.89
17	8.50	1029.50	1.59
18	9.00	1043.20	1.31
19	9.50	1054.40	1.06
20	10.00	1063.90	0.89
21	10.50	1071.50	0.71
22	11.00	1077.90	0.59

USER INPUT BLANK VALUE

BLANK VALUE = 24.63973 micrograms carbon

BLANK FACTOR = 24.63973 / 10.99988 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(1077.9 - 24.63494) (1)/(1) = +1.0533E+03 g/L Carbon
 (1077.9 - 24.63494) (1)/(1) (12) = +8.7772E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0506g

3582

LABCORE Completed Worklist Report for Worklist# 26751

Analyst: krm Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	6.02E+02	6.05E+2	100.498 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.86E+3	95.333 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	4.00E+0	4.000 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	3.80E+0	3.800 ug/g	
3 SAMPLE	S98T002711	0	@TICTOC1 TIC-02	SOLID	N/A	2.31E+04	5.000 ug/g	
3 SAMPLE	S98T002711	0	@TICTOC1 TOC-02	SOLID	N/A	6.56E+03	40.000 ug/g	
4 DUP	S98T002711	0	@TICTOC1 TIC-02	SOLID	2.31E+4	2.38E+4	2.985 RPD	
4 DUP	S98T002711	0	@TICTOC1 TOC-02	SOLID	6.56E+3	1.26E+4	63.048 RPD	
5 SAMPLE	S98T002840	0	@TICTOC1 TIC-02	SOLID	N/A	2.91E+03	5.000 ug/g	
5 SAMPLE	S98T002840	0	@TICTOC1 TOC-02	SOLID	N/A	1.73E+03	40.000 ug/g	
6 DUP	S98T002840	0	@TICTOC1 TIC-02	SOLID	2.91E+3	8.11E+3	94.374 RPD	
6 DUP	S98T002840	0	@TICTOC1 TOC-02	SOLID	1.73E+3	2.53E+3	37.559 RPD	
7 SPK	S98T002840	0	@TICTOC1 TIC-02	SOLID	1.00E+02	2.11E+02	211.000 % Recovery	
7 SPK	S98T002840	0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.43E+02	143.000 % Recovery	

Final page for worklist# 26751

Analyst Signature _____ Date _____

Analyst Signature *C. P. Chien* Date *11/25/98*

Reviewer Signature *[Signature]* Date *12/1/98*

Samples will be Rerun
Due to QC Failures. n/c

**DATA
NOT USED**

LABCORE Data Entry Template for Worklist# 26751

Analyst: KRM Instrument: CARB2 Book# 25N120
26N12FMethod: LA-342-100 Rev/Mod F-1

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@TICTOC1	SOLID		
2	BLNK				@TICTOC1	SOLID		
3	SAMPLE	S98T002711 0			@TICTOC1	SOLID	98000436	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
4	DUP	S98T002711 0			@TICTOC1	SOLID		
5	SAMPLE	S98T002737 0			@TICTOC1	SOLID	98000490	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
6	DUP	S98T002737 0			@TICTOC1	SOLID		
7	SPK	S98T002737 0			@TICTOC1	SOLID		
8	SAMPLE	S98T002840 0			@TICTOC1	SOLID	98000437	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
9	DUP	S98T002840 0			@TICTOC1	SOLID		
10	SPK	S98T002840 0			@TICTOC1	SOLID		

Final page for worklist # 26751

Analyst Signature

Date

11-15-98

Analyst Signature

Date

11/25/98

FAXED
11-15-98DATA
NOT USED

Data Entry Comments:

Batch # 5352

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3584

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	15.6	30.4
26751	µg of Carbon from Baseline (C2)	11.6	34.2
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005352			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/25/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
		5	40
Sample Point	µg of Carbon	4.00E+00	3.80E+00
BY-105			

**DATA
NOT USED**

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK WB1 REV 1.0

342100ML

3585

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	616.3	606.6
26751	µg of Carbon from Baseline (C2)	11.6	34.2
Test Code	Standard Book Number	25N12D	26N12F
@TIC/TOC1	Standard Value (µg/ml)	602	3000

Matrix
LIQUIDBatch Number
98005352Rerun
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
CJOChemist
MJLAnalyst
KRMDate Complete
11/25/98Analysis Date
11/13/98Analysis Time
12:00 AMSample Point
BY-105

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
12:00 AM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.05E+02	2.86E+03
BY-105	Percent Standard Recovery	100.4	95.4

DATA
NOT USED

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	UA	Date:	

STANDARD.WB1 REV 1.0

342100ML

3586

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.1089	0.1089
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	2530.2	748.3
26751	µg of Carbon from Baseline (C2)	11.6	34.2
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005352			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002711			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/25/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
Sample Point	µg of Carbon/g	2.31E+04	6.56E+03
BY-105			

DATA
NOT USED

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3587

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Size in g (SS)	TIC	TOC
DUP	Dilution Factor (DF)	0.1296	0.1296
Work List	µg of Carbon in Sample (C1)	1	1
26751	µg of Carbon from Baseline (C2)	3101.3	1667.9
Test Code	Known µg of C from Original Sample	11.6	34.2
@TICTOC1		2.31E+4	6.56E+3
Matrix			
SOLID			
Batch Number			
98005352			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002711			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/25/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
		5	40
Sample Point	µg of Carbon/g	2.38E+04	1.26E+04
BY-105			

DATA
NOT USED

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

WORKBOOK PAGE: SAM5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.2222	0.2222
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	657.1	418.4
26751	µg of Carbon from Baseline (C2)	11.6	34.2
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005352			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002840			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/25/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
		5	40
Sample Point	µg of Carbon/g	2.91E+03	1.73E+03
BY-105			

DATA
NOT USED

Data Entered By:

CJO

Date: 11/25/98

Signature of Chemist:

NA

Date:

SAMPLE.WB1 REV 1.0

342100ML

3589

I:\342100\OUT\26751.WB1

11/25/98

14:00:37

WORKBOOK PAGE: DUP6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.2822	0.2822
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	2301.6	747.4
26751	µg of Carbon from Baseline (C2)	11.6	34.2
Test Code	Known µg of C from Original Sample	2.91E+3	1.73E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005352			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002840			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/25/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
Sample Point	µg of Carbon/g	8.11E+03	2.53E+03
BY-105			

DATA
NOT USED

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV.2.0

<<< BLANK ANALYSIS >>>

Sample: BASE2

Date: 11/15/98

Time: 10:00:44

Sample Size = 1 ul
 Dil Factor = 1
 Blank ID # = BASE2
 Blank Value = N/A

Analyst : KR MONTEITH
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	0.90	55.56
4	1.50	1.30	47.06
5	2.00	1.80	29.17
6	2.50	2.30	20.00
7	3.00	3.60	16.67
8	3.50	4.10	12.20
9	4.00	4.70	12.77
10	4.50	5.20	9.62
11	5.00	5.70	8.77
12	5.50	6.30	9.52
13	6.00	6.90	8.70
14	6.50	7.40	6.76
15	7.00	7.80	5.13
16	7.50	8.40	7.14
17	8.00	9.00	6.67
18	8.50	9.40	4.26
19	9.00	9.90	5.05
20	9.50	10.50	5.71
21	10.00	11.00	4.55
22	10.50	11.60	5.17

DATA
NOT USED

BLANK VALUE = 11.6 micrograms carbon
 BLANK FACTOR = 11.6 / 10.49992 = 1.10E+00 ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGE 3592 TO 3607.

Sample Run By:

KR MONTEITH

00000

3592

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE2

Date: 11/15/98

Time: 10:25:22

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE2

Blank Value = N/A

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ===== Coulometer ===== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	3.20	81.25
4	1.50	12.00	73.33
5	2.00	17.80	32.58
6	2.50	21.10	15.64
7	3.00	23.10	8.66
8	3.50	24.40	5.33
9	4.00	25.50	4.31
10	4.50	26.50	3.77
11	5.00	27.40	3.28
12	5.50	28.20	2.84
13	6.00	28.90	2.42
14	6.50	29.60	2.36
15	7.00	30.20	1.99
16	7.50	31.00	2.58
17	8.00	31.50	1.59
18	8.50	32.00	1.56
19	9.00	32.70	2.14
20	9.50	33.20	1.51
21	10.00	33.70	1.48
22	10.50	34.20	1.46

DATA
NOT USED

BLANK VALUE = 34.2 micrograms carbon

BLANK FACTOR = 34.2 / 10.49817 = +3.26E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3593

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 11/15/98

Time: 10:38:33

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	6.80	89.71
4	1.50	120.00	94.33
5	2.00	289.30	58.52
6	2.50	423.80	31.74
7	3.00	509.50	16.82
8	3.50	559.40	8.92
9	4.00	586.20	4.57
10	4.50	599.20	2.17
11	5.00	604.90	0.94
12	5.50	607.80	0.48
13	6.00	609.40	0.26
14	6.50	610.70	0.21
15	7.00	611.60	0.15
16	7.50	612.50	0.15
17	8.00	613.20	0.11
18	8.50	613.80	0.10
19	9.00	614.50	0.11
20	9.50	615.10	0.10
21	10.00	615.80	0.11
22	10.50	616.30	0.08

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(616.3 - 11.54792) (1)/(1000) = +6.048E-01 g/L Carbon
 (616.3 - 11.54792) (1)/(1000) (12) = +5.040E-02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3594

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD2

Date: 11/15/98

Time: 10:49:59

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	17.80	96.63
4	1.51	198.40	91.03
5	2.00	427.80	53.62
6	2.50	530.30	19.33
7	3.00	566.90	6.46
8	3.50	582.80	2.73
9	4.00	589.90	1.20
10	4.50	593.50	0.61
11	5.00	595.70	0.37
12	5.50	597.20	0.25
13	6.00	598.40	0.20
14	6.50	599.70	0.22
15	7.00	600.80	0.18
16	7.50	601.80	0.17
17	8.00	602.60	0.13
18	8.50	603.60	0.17
19	9.00	604.40	0.13
20	9.50	605.20	0.13
21	10.00	605.70	0.08
22	10.50	606.60	0.15

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 =

+3.3E+00

ug/min Carbon

SAMPLE RESULTS:

(606.6 - 34.22343) (1)/(200) =

+2.862E+00

g/L Carbon

(606.6 - 34.22343) (1)/(200) (12) =

+2.385E-01

Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3595

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK2

Date: 11/15/98

Time: 11:01:39

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	% Difference
1	0.08		0.00	0.00
2	0.51		0.80	100.00
3	1.01		1.70	52.94
4	1.51		4.00	57.50
5	2.00		5.20	23.08
6	2.50		6.10	14.75
7	3.00		6.90	11.59
8	3.50		7.60	9.21
9	4.00		8.20	7.32
10	4.50		8.70	5.75
11	5.00		9.30	6.45
12	5.50		9.90	6.06
13	6.00		10.50	5.71
14	6.50		11.10	5.41
15	7.00		11.60	4.31
16	7.50		12.20	4.92
17	8.00		12.80	4.69
18	8.50		13.30	3.76
19	9.00		13.80	3.62
20	9.50		14.40	4.17
21	10.00		15.00	4.00
22	10.50		15.60	3.85

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817

+1.1E+00

ug/min Carbon

SAMPLE RESULTS:

(15.6 - 11.54893) (1)/(1) =

+4.05E+00

g/L Carbon

(15.6 - 11.54893) (1)/(1) (12) =

+3.38E-01

Molar Carbon

Sample Run By:

KR MONTEITH

00000

3596

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK2

Date: 11/15/98

Time: 11:13:37

Sample Size = 1 uL

Analyst : KR MONTEITH

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 3.26 ug/minute C

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	3.10	83.87
4	1.50	10.20	69.61
5	2.00	15.30	33.33
6	2.50	18.10	15.47
7	3.00	19.90	9.05
8	3.50	20.90	4.78
9	4.00	21.90	4.57
10	4.50	22.70	3.52
11	5.00	23.50	3.40
12	5.50	24.10	2.49
13	6.00	24.80	2.82
14	6.50	25.30	1.98
15	7.00	26.00	2.69
16	7.50	26.70	2.62
17	8.00	27.40	2.55
18	8.50	28.00	2.14
19	9.00	28.60	2.10
20	9.50	29.20	2.05
21	10.00	29.80	2.01
22	10.50	30.40	1.97

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(30.4 - 34.22343) (1)/(1) = < 5.00 E-3 g/L Carbon
 (30.4 - 34.22343) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3597

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2711-2

Date: 11/15/98

Time: 11:31:07

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C.

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.08	0.00	0.00
2	0.51	1.00	100.00
3	1.01	3.70	72.97
4	1.50	173.10	97.86
5	2.00	823.90	78.99
6	2.50	1570.80	47.55
7	3.00	2093.70	24.97
8	3.50	2299.20	8.94
9	4.00	2403.20	4.33
10	4.50	2462.20	2.40
11	5.00	2491.10	1.16
12	5.50	2504.00	0.52
13	6.00	2510.50	0.26
14	6.50	2514.40	0.16
15	7.00	2517.50	0.12
16	7.50	2519.80	0.09
17	8.00	2521.80	0.08
18	8.50	2523.80	0.08
19	9.00	2525.50	0.07
20	9.50	2527.10	0.06
21	10.00	2528.80	0.07
22	10.50	2530.20	0.06

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817 =

+1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(2530.2 - 11.54893) (1)/(1) =

+2.5187E+03 g/L Carbon

(2530.2 - 11.54893) (1)/(1) (12) =

+2.0989E+02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3598

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2711-2

Date: 11/15/98

Time: 11:42:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.54	1.40	100.00
3	1.04	32.20	95.65
4	1.54	291.90	88.97
5	2.04	496.40	41.20
6	2.54	595.00	16.57
7	3.04	649.10	8.33
8	3.54	680.90	4.67
9	4.04	698.80	2.56
10	4.54	709.80	1.55
11	5.04	717.10	1.02
12	5.54	722.90	0.80
13	6.04	727.20	0.59
14	6.54	731.00	0.52
15	7.04	734.40	0.46
16	7.53	737.00	0.35
17	8.03	739.50	0.34
18	8.53	741.60	0.28
19	9.03	743.60	0.27
20	9.53	745.20	0.21
21	10.03	746.90	0.23
22	10.53	748.30	0.19

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(748.3 - 34.33705) (1)/(1) = +7.140E+02 g/L Carbon
 (748.3 - 34.33705) (1)/(1) (12) = +5.950E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3599

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

Sample: S98T2711DUP-2

Date: 11/15/98

Time: 11:56:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.08	0.00	0.00
2	0.51	1.20	100.00
3	1.01	3.70	67.57
4	1.51	201.30	98.16
5	2.00	895.50	77.52
6	2.50	1642.20	45.47
7	3.00	2390.30	31.30
8	3.50	2856.30	16.31
9	4.00	2967.00	3.73
10	4.50	3027.80	2.01
11	5.00	3057.00	0.96
12	5.50	3070.20	0.43
13	6.00	3077.10	0.22
14	6.50	3081.90	0.16
15	7.00	3085.30	0.11
16	7.50	3088.30	0.10
17	8.00	3090.80	0.08
18	8.50	3093.20	0.08
19	9.00	3095.60	0.08
20	9.50	3097.50	0.06
21	10.00	3099.60	0.07
22	10.50	3101.30	0.05

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(3101.3 - 11.54879) (1)/(1) =

+3.0898E+03 g/L Carbon

(3101.3 - 11.54879) (1)/(1) (12) =

+2.5748E+02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3600

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2711DUP-2

Date: 11/15/98

Time: 12:14:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	2.10	100.00
3	1.01	43.50	95.17
4	1.51	579.60	92.49
5	2.00	1146.40	49.44
6	2.50	1379.10	16.87
7	3.00	1479.50	6.79
8	3.50	1539.10	3.87
9	4.00	1574.70	2.26
10	4.50	1596.40	1.36
11	5.00	1610.80	0.88
12	5.50	1620.90	0.64
13	6.00	1629.00	0.50
14	6.50	1635.60	0.40
15	7.00	1641.40	0.35
16	7.50	1646.20	0.29
17	8.00	1650.90	0.28
18	8.50	1654.90	0.24
19	9.00	1658.50	0.22
20	9.50	1662.00	0.21
21	10.00	1665.10	0.19
22	10.50	1667.90	0.17

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817

+3.3E+00

ug/min Carbon

SAMPLE RESULTS:

(1667.9 - 34.22662) (1)/(1) =

+1.6337E+03 g/L Carbon

(1667.9 - 34.22662) (1)/(1) (12) =

+1.3614E+02 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3601

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV: 2.0

Sample: S98T2840-2

Date: 11/15/98

Time: 14:23:44

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.00	2.50	68.00
4	1.50	26.50	90.57
5	2.00	135.20	80.40
6	2.50	303.20	55.41
7	3.00	458.90	33.93
8	3.50	547.60	16.20
9	4.00	598.40	8.49
10	4.50	623.50	4.03
11	5.00	634.40	1.72
12	5.50	639.50	0.80
13	6.00	642.70	0.50
14	6.50	645.00	0.36
15	7.00	646.90	0.29
16	7.50	648.50	0.25
17	8.00	650.30	0.28
18	8.50	651.70	0.21
19	9.00	653.20	0.23
20	9.50	654.60	0.21
21	10.00	655.90	0.20
22	10.50	657.10	0.18

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817

+1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(657.1 - 11.54785) (1)/(1) =

+6.456E+02 g/L Carbon

(657.1 - 11.54785) (1)/(1) (12) =

+5.380E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3602

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2840-2

Date: 11/15/98

Time: 14:35:32

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute.C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference
1	0.08		0.00	0.00
2	0.51		1.60	100.00
3	1.01		11.10	85.59
4	1.50		112.70	90.15
5	2.00		228.90	50.76
6	2.50		294.50	22.28
7	3.00		332.80	11.51
8	3.50		354.80	6.20
9	4.00		368.70	3.77
10	4.50		378.10	2.49
11	5.00		385.30	1.87
12	5.50		391.00	1.46
13	6.00		395.40	1.11
14	6.50		399.30	0.98
15	7.00		402.40	0.77
16	7.50		405.60	0.79
17	8.00		408.00	0.59
18	8.50		410.60	0.63
19	9.00		412.70	0.51
20	9.50		414.80	0.51
21	10.00		416.90	0.50
22	10.50		418.40	0.36

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(418.4 - 34.22343) (1)/(1) = +3.842E+02 g/L Carbon
 (418.4 - 34.22343) (1)/(1) (12) = +3.201E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3603

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2840DUP-2

Date: 11/15/98

Time: 14:47:04

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.80	100.00
3	1.01	4.40	59.09
4	1.51	120.60	96.35
5	2.00	485.90	75.18
6	2.50	1168.00	58.40
7	3.00	1822.70	35.92
8	3.50	2073.00	12.07
9	4.00	2177.50	4.80
10	4.50	2234.00	2.53
11	5.00	2261.20	1.20
12	5.50	2273.20	0.53
13	6.00	2279.10	0.26
14	6.50	2283.60	0.20
15	7.00	2286.40	0.12
16	7.50	2289.40	0.13
17	8.00	2291.80	0.10
18	8.50	2294.00	0.10
19	9.00	2295.90	0.08
20	9.50	2298.00	0.09
21	10.00	2299.70	0.07
22	10.50	2301.60	0.08

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(2301.6 - 11.54893) (1)/(1) = +2.2901E+03 g/L Carbon
 (2301.6 - 11.54893) (1)/(1) (12) = +1.9084E+02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3604

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2840DUP-2

Date: 11/15/98

Time: 15:21:53

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	2.00	100.00
3	1.01	16.10	87.58
4	1.50	215.70	92.54
5	2.00	424.00	49.13
6	2.50	541.90	21.76
7	3.00	607.70	10.83
8	3.50	647.60	6.16
9	4.00	671.30	3.53
10	4.50	686.80	2.26
11	5.00	697.60	1.55
12	5.50	706.10	1.20
13	6.00	712.90	0.95
14	6.50	718.60	0.79
15	7.00	723.60	0.69
16	7.50	728.10	0.62
17	8.00	732.10	0.55
18	8.50	735.60	0.48
19	9.00	738.90	0.45
20	9.50	742.00	0.42
21	10.00	744.70	0.36
22	10.50	747.40	0.36

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(747.4 - 34.23) (1)/(1) = +7.132E+02 g/L Carbon
 (747.4 - 34.23) (1)/(1) (12) = +5.943E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3605

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2840SPK-2

Date: 11/15/98

Time: 15:40:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.1 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	1.20	100.00
3	1.01	4.70	74.47
4	1.50	138.50	96.61
5	2.00	441.50	68.63
6	2.50	706.60	37.52
7	3.00	889.20	20.54
8	3.50	995.60	10.69
9	4.00	1062.10	6.26
10	4.50	1099.80	3.43
11	5.00	1117.90	1.62
12	5.50	1125.60	0.68
13	6.00	1129.40	0.34
14	6.50	1132.70	0.29
15	7.00	1134.70	0.18
16	7.50	1137.00	0.20
17	8.00	1138.80	0.16
18	8.50	1140.40	0.14
19	9.00	1142.30	0.17
20	9.50	1143.90	0.14
21	10.00	1145.40	0.13
22	10.50	1146.80	0.12

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 11.54799 micrograms carbon

BLANK FACTOR = 11.54799 / 10.49817 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(1146.8 - 11.54785) (1)/(1) =

+1.1353E+03 g/L Carbon

(1146.8 - 11.54785) (1)/(1) (12) =

+9.4604E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3606

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV.2:0

Sample: S98T2840SPK-2

Date: 11/15/98

Time: 15:53:00

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.26 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	1.50	100.00
3	1.01	25.30	94.07
4	1.51	214.50	88.21
5	2.00	436.70	50.88
6	2.50	567.30	23.02
7	3.00	634.50	10.59
8	3.50	673.90	5.85
9	4.00	696.10	3.19
10	4.50	709.50	1.89
11	5.00	718.10	1.20
12	5.50	724.90	0.94
13	6.00	730.80	0.81
14	6.50	735.30	0.61
15	7.00	739.70	0.59
16	7.50	743.30	0.48
17	8.00	746.60	0.44
18	8.50	749.50	0.39
19	9.03	752.60	0.41
20	9.53	755.10	0.33
21	10.03	757.80	0.36
22	10.53	760.00	0.29

DATA
NOT USED

USER INPUT BLANK VALUE

BLANK VALUE = 34.22403 micrograms carbon

BLANK FACTOR = 34.22403 / 10.49817 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(760 - 34.33685) (1)/(1) = +7.257E+02 g/L Carbon
 (760 - 34.33685) (1)/(1) (12) = +6.047E+01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

KR MONTEITH

00000

3607

LABCORE Completed Worklist Report for Worklist# 26752

Analyst: slh Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@TICTOC1 TIC-02	SOLID	6.02E+02	5.77E+2	95.847 % Recovery	
1 STD	0	@TICTOC1 TOC-02	SOLID	3.00E+03	2.87E+3	95.667 % Recovery	
2 BLNK	0	@TICTOC1 TIC-02	SOLID	1	5.90E+0	5.900 ug/g	
2 BLNK	0	@TICTOC1 TOC-02	SOLID	1	5.20E+0	5.200 ug/g	
3 SAMPLE	S98T002841 0	@TICTOC1 TIC-02	SOLID	N/A	4.39E+03	5.000 ug/g	
3 SAMPLE	S98T002841 0	@TICTOC1 TOC-02	SOLID	N/A	8.24E+03	40.000 ug/g	
4 DUP	S98T002841 0	@TICTOC1 TIC-02	SOLID	4.39E+3	4.58E+3	4.236 RPD	
4 DUP	S98T002841 0	@TICTOC1 TOC-02	SOLID	8.24E+3	1.10E+4	28.690 RPD	
5 TRIPL	S98T002841 0	@TICTOC1 TIC-02	SOLID	4.39E+3	5.04E+3	13.786 RPD	
5 TRIPL	S98T002841 0	@TICTOC1 TOC-02	SOLID	8.24E+3	1.17E+4	34.704 RPD	
6 SAMPLE	S98T002910 0	@TICTOC1 TIC-02	SOLID	N/A	1.90E+03	5.000 ug/g	
6 SAMPLE	S98T002910 0	@TICTOC1 TOC-02	SOLID	N/A	1.69E+3	40.000 ug/g	
7 DUP	S98T002910 0	@TICTOC1 TIC-02	SOLID	1.90E+3	8.88E+2	72.597 RPD	
7 DUP	S98T002910 0	@TICTOC1 TOC-02	SOLID	1.69E+3	1.58E+3	6.728 RPD	
8 TRIPL	S98T002910 0	@TICTOC1 TIC-02	SOLID	1.90E+3	1.53E+3	21.574 RPD	
8 TRIPL	S98T002910 0	@TICTOC1 TOC-02	SOLID	1.69E+3	<4.00E+1	RPD	
9 SAMPLE	S98T002911 0	@TICTOC1 TIC-02	SOLID	N/A	3.03E+03	5.000 ug/g	
9 SAMPLE	S98T002911 0	@TICTOC1 TOC-02	SOLID	N/A	1.14E+03	40.000 ug/g	
10 DUP	S98T002911 0	@TICTOC1 TIC-02	SOLID	3.03E+3	3.13E+3	3.247 RPD	
10 DUP	S98T002911 0	@TICTOC1 TOC-02	SOLID	1.14E+3	1.11E+3	2.667 RPD	
11 SPK	S98T002911 0	@TICTOC1 TIC-02	SOLID	1.00E+02	1.12E+02	112.000 % Recovery	
11 SPK	S98T002911 0	@TICTOC1 TOC-02	SOLID	1.00E+02	9.21E+01	92.100 % Recovery	

Final page for worklist# 26752

Analyst Signature _____ Date _____

Analyst Signature *slh* 12/1/98 Date

Reviewer Signature *slh* 12/2/98 Date

LABCORE Data Entry Template for Worklist# 26752

Analyst: SK Instrument: CARB2 Book# TK 25N12-D
 Method: LA-342-100 Rev/Mod E-1 TOC 26N12-F
 Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A Test	Matrix	Group#	Project
1 STD			@TICTOC1 SOLID		
2 BLNK			@TICTOC1 SOLID		
3 SAMPLE	S98T002841 0	@TICTOC1 SOLID		98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02				
4 DUP	S98T002841 0	@TICTOC1 SOLID			
5 SAMPLE	S98T002910 0	@TICTOC1 SOLID		98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02				
6 DUP	S98T002910 0	@TICTOC1 SOLID			
7 SAMPLE	S98T002911 0	@TICTOC1 SOLID		98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02				
8 DUP	S98T002911 0	@TICTOC1 SOLID			
9 SPK	S98T002911 0	@TICTOC1 SOLID			

Final page for worklist # 26752

Analyst Signature James L. Hood Date 11-27-98

Analyst Signature Q. Quinn Date 12/1/98

Data Entry Comments: Batch # 5353

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	583.5	594.5
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000
Matrix			
LIQUID			
Batch Number			
98005353			
Rerun			
0	QC Actual in µg/mL = Standard Value (µg/mL)		
Sample Prep	QC Found in µg/mL = (C1 - C2) * DF / SS		
N/A	QC Found in µg/mL for TIC = 5 if C1 < C2		
Sample #	QC Found in µg/mL for TOC = 40 if C1 < C2		
STD			
Instrument Code	% Recovery = QC Found / QC Actual * 100		
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
10:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.77E+02	2.87E+03
BY-105	Percent Standard Recovery	95.8	95.5

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	
STANDARD.WB1 REV 1.0	342100ML		

3610

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	12.5	26.5
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	5	40
10:00 PM			
Sample Point	µg of Carbon	5.90E+00	5.20E+00
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3611

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0588	0.0588
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	265	505.6
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002841			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
10:00 PM		5	40
Sample Point	µg of Carbon/g	4.39E+03	8.24E+03
BY-105			

Data Entered By:

CJO

Date:

12/01/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0651	0.0651
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	304.9	734.2
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Known µg of C from Original Sample	4.39E+3	8.24E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002841			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
10:00 PM		5	40
Sample Point	µg of Carbon/g	4.58E+03	1.10E+04
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3613

WORKBOOK PAGE: TRIPL5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0806	0.0806
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	412.8	965.8
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Known µg of C from Original Sample	4.39E+3	8.24E+3
@TICTOC1	Duplicate Result	4.58E+3	1.10E+4
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002841			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
10:00 PM		5	40
Sample Point	µg of Carbon/g	5.04E+03	1.17E+04
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3614

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0737	0.0737
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	146.3	145.8
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code			
@TICTOC1			

Matrix
SOLID

Batch Number
98005353

Rerun
0

µg of Carbon/g = (C1-C2) * DF / SS
 µg of Carbon/g for TIC = 5 if C1 < C2
 µg of Carbon/g for TOC = 40 if C1 < C2

Sample Prep
N/A

Sample #
S98T002910

Instrument Code
CARB2

Prepared By
CJO

Chemist
MJL

Analyst
SLH

Date Complete
12/01/98

Analysis Date
11/27/98

Analysis Time
10:00 PM

Method Detection Limit in ug/g

Sample Point
BY-105

µg of Carbon/g

TIC	TOC
5	40
1.90E+03	1.69E+03

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0828	0.0828
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	80.1	152.4
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Known µg of C from Original Sample	1.90E+3	1.69E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002910			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
10:00 PM		5	40
Sample Point	µg of Carbon/g	8.88E+02	1.58E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3616

WORKBOOK PAGE: TRIPL8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0869	0.0869
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	139.2	0
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Known µg of C from Original Sample	1.90E+3	1.69E+3
@TICTOC1	Duplicate Result	8.88E+2	1.58E+3
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002910			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date	NOTE: FOR TOC: The Reported Result is	<	40.
11/27/98			
Analysis Time	Method Detection Limit in ug/g	5	40
10:00 PM			
Sample Point	µg of Carbon/g Note:	<	
BY-105		1.53E+03	4.00E+01

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	NA	Date:	

WORKBOOK PAGE: SAM9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0671	0.0671
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	209.9	97.9
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002911			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
10:00 PM		5	40
Sample Point	µg of Carbon/g	3.03E+03	1.14E+03
BY-105			

Data Entered By:

CJO

Date:

12/01/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0717	0.0717
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	230.9	101
26752	µg of Carbon from Baseline (C2)	6.6	21.3
Test Code	Known µg of C from Original Sample	3.03E+3	1.14E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005353			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002911			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time	Method Detection Limit in ug/g	5	40
10:00 PM			
Sample Point	µg of Carbon/g	3.13E+03	1.11E+03
BY-105			

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SPIKE11

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0671	0.0671
Worklist	Final Coulometer Reading in μg (C1)	209.9	97.9
26752	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0698	0.0698
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in μg (C2)	555	377.3
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in $\mu\text{g}/\text{mL}$ (SPK CONC)	602	3000
98005353	μg C in baseline (BL)	6.6	21.3
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = $((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100$		
S98T002911			
Instrument Code	QC Actual in $\mu\text{g}/\text{mL}$ = Spike Value ($\mu\text{g}/\text{mL}$)		
CARB2	QC Found in $\mu\text{g}/\text{mL}$ = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/01/98			
Analysis Date			
11/27/98			
Analysis Time			
10:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.74E+02	2.76E+03
BY-105	Percent Spike Recovery	111.9	92.1

Data Entered By:	CJO	Date:	12/01/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3620

HNF-1663 REV 0

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 11/26/98

Time: 19:29:54

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	1.30	0.00
2	1.01	2.80	53.57
3	1.51	4.10	31.71
4	2.01	5.40	24.07
5	2.51	19.40	72.16
6	3.00	74.80	74.06
7	3.50	159.20	53.02
8	4.00	251.60	36.72
9	4.50	338.30	25.63
10	5.00	413.50	18.19
11	5.50	471.50	12.30
12	6.00	513.20	8.13
13	6.50	540.70	5.09
14	7.00	557.40	3.00
15	7.50	567.10	1.71
16	8.00	573.10	1.05
17	8.50	576.70	0.62
18	9.00	579.30	0.45
19	9.50	580.70	0.24
20	10.00	582.20	0.26
21	10.50	582.90	0.12
22	11.00	583.50	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(583.5 - 6.599341) (1)/(1000) = +5.769E-01 g/L Carbon
(583.5 - 6.599341) (1)/(1000) (12) = +4.808E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 3621 TO 3642.

Sample Run By:

SL HOOD

00002

1ml = 25N/12.0

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 11/26/98

Time: 19:42:05

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.80	62.50
3	1.50	1.30	38.46
4	2.01	1.90	31.58
5	2.51	5.80	67.24
6	3.01	33.00	82.42
7	3.51	113.50	70.93
8	4.00	242.60	53.22
9	4.50	366.90	33.88
10	5.00	456.10	19.56
11	5.50	512.80	11.06
12	6.00	545.70	6.03
13	6.50	564.50	3.33
14	7.00	575.10	1.84
15	7.50	581.70	1.13
16	8.00	585.20	0.60
17	8.50	588.20	0.51
18	9.00	590.10	0.32
19	9.50	591.40	0.22
20	10.00	592.70	0.22
21	10.50	593.60	0.15
22	11.00	594.50	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(594.5 - 21.33929) (1)/(200) = +2.866E+00 g/L Carbon
 (594.5 - 21.33929) (1)/(200) (12) = +2.388E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

200ml 26N12-F

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/26/98

Time: 19:04:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.10	0.00
2	1.01	0.30	66.67
3	1.51	0.50	40.00
4	2.00	0.70	28.57
5	2.50	1.20	41.67
6	3.00	2.10	42.86
7	3.50	2.60	19.23
8	4.00	3.10	16.13
9	4.50	3.60	13.89
10	5.00	3.90	7.69
11	5.50	4.20	7.14
12	6.00	4.50	6.67
13	6.50	4.70	4.26
14	7.00	5.00	6.00
15	7.50	5.20	3.85
16	8.00	5.40	3.70
17	8.50	5.60	3.57
18	9.00	5.90	5.08
19	9.50	6.10	3.28
20	10.00	6.20	1.61
21	10.50	6.40	3.13
22	11.00	6.60	3.03

BLANK VALUE = 6.6 micrograms carbon

BLANK FACTOR = 6.6 / 10.99707 =

+6.0E-01

ug/min Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 11/26/98

Time: 19:16:43

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.10	0.00
2	1.01	0.30	66.67
3	1.50	0.50	40.00
4	2.00	0.70	28.57
5	2.50	1.20	41.67
6	3.00	2.20	45.45
7	3.50	3.50	37.14
8	4.00	5.20	32.69
9	4.50	6.30	17.46
10	5.00	7.50	16.00
11	5.50	8.30	9.64
12	6.00	9.10	8.79
13	6.50	9.70	6.19
14	7.00	10.50	7.62
15	7.50	11.20	6.25
16	8.00	12.10	7.44
17	8.50	13.30	9.02
18	9.00	14.60	8.90
19	9.50	16.10	9.32
20	10.00	18.00	10.56
21	10.50	19.70	8.63
22	11.00	21.30	7.51

BLANK VALUE = 21.3 micrograms carbon

BLANK FACTOR = 21.3 / 10.9978 = +1.94E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/26/98

Time: 19:54:14

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	=== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.20	50.00
3	1.51	1.80	33.33
4	2.00	2.30	21.74
5	2.50	3.40	32.35
6	3.00	4.60	26.09
7	3.50	5.60	17.86
8	4.00	6.60	15.15
9	4.50	7.30	9.59
10	5.00	7.90	7.59
11	5.50	8.40	5.95
12	6.00	8.90	5.62
13	6.50	9.30	4.30
14	7.00	9.70	4.12
15	7.50	10.10	3.96
16	8.00	10.50	3.81
17	8.50	10.90	3.67
18	9.00	11.20	2.68
19	9.50	11.50	2.61
20	10.00	11.80	2.54
21	10.50	12.20	3.28
22	11.00	12.50	2.40

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(12.5 - 6.598608) (1)/(1) = +5.90E+00 g/L Carbon
 (12.5 - 6.598608) (1)/(1) (12) = +4.92E-01 Molar Carbon

Sample Run By:

SL HOOD

00002

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 11/26/98

Time: 22:38:59

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.60	33.33
3	1.50	1.10	45.45
4	2.00	1.30	15.38
5	2.50	2.70	51.85
6	3.00	5.50	50.91
7	3.50	9.50	42.11
8	4.00	13.40	29.10
9	4.50	16.50	18.79
10	5.00	18.60	11.29
11	5.50	20.10	7.46
12	6.00	20.90	3.83
13	6.50	21.80	4.13
14	7.00	22.50	3.11
15	7.50	23.00	2.17
16	8.00	23.50	2.13
17	8.50	24.10	2.49
18	9.00	24.50	1.63
19	9.50	24.90	1.61
20	10.00	25.40	1.97
21	10.50	26.10	2.68
22	11.00	26.50	1.51

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(26.5 - 21.33598) (1) / (1) = +5.16E+00 g/L Carbon
 (26.5 - 21.33598) (1) / (1) (12) = +4.30E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T002841

Date: 11/27/98

Time: 00:25:37

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	2.20	0.00
2	1.00	3.40	35.29
3	1.50	4.20	19.05
4	2.00	5.40	22.22
5	2.50	8.50	36.47
6	3.00	22.10	61.54
7	3.50	52.40	57.82
8	4.00	92.80	43.53
9	4.50	135.40	31.46
10	5.00	173.80	22.09
11	5.50	204.40	14.97
12	6.00	226.10	9.60
13	6.50	240.00	5.79
14	7.00	248.60	3.46
15	7.50	254.00	2.13
16	8.00	257.20	1.24
17	8.50	259.20	0.77
18	9.00	260.80	0.61
19	9.50	262.10	0.50
20	10.00	263.20	0.42
21	10.50	264.30	0.42
22	11.00	265.00	0.26

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 =

+6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(265 - 6.597701) (1)/(1) =

+2.584E+02 g/L Carbon

(265 - 6.597701) (1)/(1) (12) =

+2.153E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0588g

3627

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S09T002841

Date: 11/27/98

Time: 00:38:21

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.20	41.67
3	1.50	2.30	47.83
4	2.00	21.40	89.25
5	2.50	78.50	72.74
6	3.00	149.40	47.46
7	3.50	218.50	31.62
8	4.00	278.30	21.49
9	4.50	325.40	14.47
10	5.00	362.10	10.14
11	5.50	389.80	7.11
12	6.00	412.30	5.46
13	6.50	430.50	4.23
14	7.00	445.40	3.35
15	7.50	457.70	2.69
16	8.00	467.80	2.16
17	8.50	476.30	1.78
18	9.00	484.20	1.63
19	9.50	490.40	1.26
20	10.00	496.10	1.15
21	10.50	501.30	1.04
22	11.00	505.60	0.85

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(505.6 - 21.33579) (1)/(1) = +4.843E+02 g/L Carbon
 (505.6 - 21.33579) (1)/(1) (12) = +4.036E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0588g

3628

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02841 DUP

Date: 11/27/98

Time: 00:51:07

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ===== Coulometer ===== % Difference ==

1	0.51	2.50	0.00
2	1.00	4.80	47.92
3	1.50	6.80	29.41
4	2.00	8.80	22.73
5	2.50	18.90	53.44
6	3.00	47.80	60.46
7	3.50	97.00	50.72
8	4.00	154.20	37.09
9	4.50	204.30	24.52
10	5.00	241.70	15.47
11	5.50	265.90	9.10
12	6.00	280.40	5.17
13	6.50	288.30	2.74
14	7.00	293.10	1.64
15	7.50	296.30	1.08
16	8.00	298.40	0.70
17	8.50	300.20	0.60
18	9.00	301.20	0.33
19	9.50	302.20	0.33
20	10.00	303.10	0.30
21	10.50	304.00	0.30
22	11.00	304.90	0.30

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(304.9 - 6.599199) (1)/(1) = +2.983E+02 g/L Carbon
 (304.9 - 6.599199) (1)/(1) (12) = +2.486E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0651g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02841 DUP

Date: 11/27/98

Time: 01:04:57

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	1.00	0.00
2	1.01	1.30	23.08
3	1.51	2.40	45.83
4	2.00	20.90	88.52
5	2.50	82.40	74.64
6	3.00	170.60	51.70
7	3.50	269.70	36.74
8	4.00	363.00	25.70
9	4.50	441.60	17.80
10	5.00	504.30	12.43
11	5.50	553.80	8.94
12	6.00	592.50	6.53
13	6.50	622.90	4.88
14	7.00	646.70	3.68
15	7.50	665.90	2.88
16	8.00	681.50	2.29
17	8.50	694.30	1.84
18	9.00	704.90	1.50
19	9.50	714.40	1.33
20	10.00	722.20	1.08
21	10.50	728.20	0.82
22	11.00	734.20	0.82

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon.

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon.

SAMPLE RESULTS:

(734.2 - 21.33612) (1)/(1) = +7.129E+02 g/L Carbon
 (734.2 - 21.33612) (1)/(1) (12) = +5.941E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0651g

3630

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02841 TRP

Date: 11/27/98

Time: 01:23:02

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.80	0.00
2	1.01	1.40	42.86
3	1.50	2.00	30.00
4	2.00	3.00	33.33
5	2.50	14.80	79.73
6	3.00	59.60	75.17
7	3.50	131.70	54.75
8	4.00	207.50	36.53
9	4.50	273.60	24.16
10	5.00	324.00	15.56
11	5.50	358.70	9.67
12	6.00	379.70	5.53
13	6.50	391.50	3.01
14	7.00	398.30	1.71
15	7.50	402.60	1.07
16	8.00	404.90	0.57
17	8.50	406.90	0.49
18	9.00	408.40	0.37
19	9.50	409.50	0.27
20	10.00	410.70	0.29
21	10.50	411.70	0.24
22	11.00	412.80	0.27

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 =

+6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(412.8 - 6.598201) (1)/(1) =

+4.062E+02 g/L Carbon

(412.8 - 6.598201) (1)/(1) (12) =

+3.385E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0806g

3631

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02841 TRP

Date: 11/27/98

Time: 01:35:34

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	1.10	0.00
2	1.00	1.90	42.11
3	1.50	2.80	32.14
4	2.00	15.70	82.17
5	2.50	62.40	74.84
6	3.00	145.30	57.05
7	3.50	266.40	45.46
8	4.00	398.30	33.12
9	4.50	512.30	22.25
10	5.00	605.10	15.34
11	5.50	679.20	10.91
12	6.00	738.00	7.97
13	6.50	786.00	6.11
14	7.00	824.10	4.62
15	7.50	855.10	3.63
16	8.00	880.60	2.90
17	8.50	901.50	2.32
18	9.00	919.00	1.90
19	9.50	933.90	1.60
20	10.00	946.60	1.34
21	10.50	956.90	1.08
22	11.00	965.80	0.92

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(965.8 - 21.33417) (1)/(1) = +9.445E+02 g/L Carbon
 (965.8 - 21.33417) (1)/(1) (12) = +7.871E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = 0.0806g

HNF-1663 REV. 0

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002910

Date: 11/27/98

Time: 01:49:10

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	5.00	0.00
2	1.01	8.50	41.18
3	1.50	11.40	25.44
4	2.00	14.80	22.97
5	2.50	26.90	44.98
6	3.00	50.60	46.84
7	3.50	79.20	36.11
8	4.00	102.70	22.88
9	4.50	118.50	13.33
10	5.00	128.30	7.64
11	5.50	133.60	3.97
12	6.00	137.00	2.48
13	6.50	139.30	1.65
14	7.00	140.70	1.00
15	7.50	141.60	0.64
16	8.00	142.40	0.56
17	8.50	143.30	0.63
18	9.00	143.90	0.42
19	9.50	144.70	0.55
20	10.00	145.10	0.28
21	10.50	145.80	0.48
22	11.00	146.30	0.34

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 =

+6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(146.3 - 6.5987) (1)/(1) =

+1.397E+02 g/L Carbon

(146.3 - 6.5987) (1)/(1) (12) =

+1.164E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0737g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002910

Date: 11/27/98

Time: 02:01:21

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.10	45.45
3	1.50	1.80	38.89
4	2.00	11.50	84.35
5	2.50	39.70	71.03
6	3.00	71.40	44.40
7	3.50	95.50	25.24
8	4.00	111.00	13.96
9	4.50	120.00	7.50
10	5.00	126.00	4.76
11	5.50	129.50	2.70
12	6.00	132.50	2.26
13	6.50	135.00	1.85
14	7.00	136.90	1.39
15	7.50	138.50	1.16
16	8.00	139.80	0.93
17	8.50	141.00	0.85
18	9.00	141.90	0.63
19	9.50	143.00	0.77
20	10.00	143.90	0.63
21	10.50	144.90	0.69
22	11.00	145.80	0.62

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(145.8 - 21.33612) (1)/(1) = +1.245E+02 g/L Carbon
 (145.8 - 21.33612) (1)/(1) (12) = +1.037E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0737g

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02910 DUP Date: 11/27/98 Time: 02:13:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.80	0.00
2	1.01	1.50	46.67
3	1.51	2.20	31.82
4	2.01	3.50	37.14
5	2.50	7.40	52.70
6	3.00	17.30	57.23
7	3.50	31.30	44.73
8	4.00	45.00	30.44
9	4.50	56.20	19.93
10	5.00	63.90	12.05
11	5.50	68.70	6.99
12	6.00	72.00	4.58
13	6.50	73.60	2.17
14	7.00	74.90	1.74
15	7.50	75.90	1.32
16	8.00	76.60	0.91
17	8.50	77.20	0.78
18	9.00	77.90	0.90
19	9.50	78.50	0.76
20	10.00	78.90	0.51
21	10.50	79.50	0.75
22	11.00	80.10	0.75

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 =

+6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(80.1 - 6.598805) (1)/(1) =

+7.35E+01 g/L Carbon

(80.1 - 6.598805) (1)/(1) (12) =

+6.13E+00 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0828g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T01910 DUP

Date: 11/27/98

Time: 02:26:02

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.50	1.50	40.00
4	2.00	7.40	79.73
5	2.50	31.90	76.80
6	3.00	64.30	50.39
7	3.50	93.00	30.86
8	4.00	112.90	17.63
9	4.50	125.10	9.75
10	5.00	132.30	5.44
11	5.50	136.60	3.15
12	6.00	139.70	2.22
13	6.50	142.00	1.62
14	7.00	143.60	1.11
15	7.50	145.20	1.10
16	8.00	146.30	0.75
17	8.50	147.50	0.81
18	9.00	148.60	0.74
19	9.50	149.90	0.87
20	10.00	150.70	0.53
21	10.50	151.60	0.59
22	11.00	152.40	0.52

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(152.4 - 21.33583) (1)/(1) = +1.311E+02 g/L Carbon
 (152.4 - 21.33583) (1)/(1) (12) = +1.092E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0828g

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002911

Date: 11/26/98

Time: 22:56:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.60	50.00
3	1.50	0.80	25.00
4	2.00	1.50	46.67
5	2.50	12.20	87.70
6	3.00	48.40	74.79
7	3.50	89.90	46.16
8	4.00	124.70	27.91
9	4.50	151.80	17.85
10	5.00	171.40	11.44
11	5.50	185.20	7.45
12	6.00	193.40	4.24
13	6.50	198.90	2.77
14	7.00	202.50	1.78
15	7.50	204.80	1.12
16	8.00	206.10	0.63
17	8.50	207.30	0.58
18	9.00	208.00	0.34
19	9.50	208.50	0.24
20	10.00	209.00	0.24
21	10.50	209.50	0.24
22	11.00	209.90	0.19

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(209.9 - 6.598682) (1)/(1) =

+2.033E+02 g/L Carbon

(209.9 - 6.598682) (1)/(1) (12) =

+1.694E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0671g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T002911

Date: 11/26/98

Time: 23:09:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.70	42.86
3	1.50	1.10	36.36
4	2.00	1.80	38.89
5	2.50	7.50	76.00
6	3.01	23.40	67.95
7	3.50	43.10	45.71
8	4.00	60.70	29.00
9	4.50	73.00	16.85
10	5.00	80.80	9.65
11	5.50	85.10	5.05
12	6.00	88.70	4.06
13	6.50	90.60	2.10
14	7.00	92.30	1.84
15	7.50	93.30	1.07
16	8.00	94.30	1.06
17	8.50	95.10	0.84
18	9.00	95.90	0.83
19	9.50	96.40	0.52
20	10.00	97.00	0.62
21	10.50	97.60	0.61
22	11.00	97.90	0.31

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(97.9 - 21.33621) (1)/(1) = +7.66E+01 g/L Carbon
 (97.9 - 21.33621) (1)/(1) (12) = +6.38E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = 0.0671g

3638

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02911 DUP

Date: 11/26/98

Time: 23:26:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.20	66.67
3	1.50	1.50	20.00
4	2.00	2.10	28.57
5	2.50	9.30	77.42
6	3.00	36.50	74.52
7	3.50	76.20	52.10
8	4.00	117.00	34.87
9	4.50	152.10	23.08
10	5.00	179.30	15.17
11	5.50	198.40	9.63
12	6.00	210.50	5.75
13	6.50	217.90	3.40
14	7.00	222.70	2.16
15	7.50	224.90	0.98
16	8.00	226.80	0.84
17	8.50	228.00	0.53
18	9.00	228.90	0.39
19	9.50	229.50	0.26
20	10.00	230.00	0.22
21	10.50	230.70	0.30
22	11.00	230.90	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(230.9 - 6.598828) (1)/(1) =

+2.243E+02 g/L Carbon

(230.9 - 6.598828) (1)/(1) (12) =

+1.869E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS- .0717g

3639

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02922 DUP

Date: 11/26/98

Time: 23:39:13

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.94 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	0.80	62.50
3	1.50	1.30	38.46
4	2.00	2.10	38.10
5	2.50	11.00	80.91
6	3.00	28.80	61.81
7	3.50	48.90	41.10
8	4.00	64.80	24.54
9	4.50	76.10	14.85
10	5.00	83.20	8.53
11	5.50	87.90	5.35
12	6.00	91.00	3.41
13	6.50	92.80	1.94
14	7.00	94.40	1.69
15	7.50	95.60	1.26
16	8.00	96.50	0.93
17	8.50	97.50	1.03
18	9.00	98.50	1.02
19	9.50	99.40	0.91
20	10.00	99.90	0.50
21	10.50	100.40	0.50
22	11.00	101.00	0.59

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(101 - 21.33266) (1) / (1) = +7.967E+01 g/L Carbon
 (101 - 21.33266) (1) / (1) (12) = +6.639E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS=0.0717g

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02911 SPK

Date: 11/26/98

Time: 23:52:29

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .6 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.51	1.40	35.71
4	2.00	2.30	39.13
5	2.50	20.80	88.94
6	3.00	82.60	74.82
7	3.50	174.60	52.69
8	4.00	271.90	35.79
9	4.50	358.30	24.11
10	5.00	426.50	15.99
11	5.50	475.00	10.21
12	6.00	507.30	6.37
13	6.50	526.60	3.67
14	7.00	537.70	2.06
15	7.50	544.50	1.25
16	8.00	548.10	0.66
17	8.50	550.20	0.38
18	9.00	551.90	0.31
19	9.50	553.00	0.20
20	10.00	553.70	0.13
21	10.50	554.40	0.13
22	11.00	555.00	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 6.598682 micrograms carbon.

BLANK FACTOR = 6.598682 / 10.9978 = +6.0E-01 ug/min Carbon

SAMPLE RESULTS:

(555 - 6.599268) (1)/(1) = +5.484E+02 g/L Carbon
 (555 - 6.599268) (1)/(1) (12) = +4.570E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0698g + .500ml spk 25/12-D

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02911 ^{SPK mL} ~~DUP~~ Date: 11/27/98 Time: 00:05:22
11/20/98

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 1.94 ug/minute C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.10	63.64
3	1.51	1.40	21.43
4	2.00	2.90	51.72
5	2.50	16.30	82.21
6	3.00	55.40	70.58
7	3.50	126.90	56.34
8	4.00	206.60	38.58
9	4.50	272.90	24.29
10	5.00	315.80	13.58
11	5.50	340.90	7.36
12	6.00	354.40	3.81
13	%-1433.50	362.00	2.10
14	%-1433.00	366.50	1.23
15	%-1432.50	369.30	0.76
16	%-1432.00	371.50	0.59
17	%-1431.50	372.80	0.35
18	%-1431.00	373.90	0.29
19	%-1430.50	375.10	0.32
20	%-1430.00	375.80	0.19
21	%-1429.50	376.60	0.21
22	%-1429.01	377.30	0.19

USER INPUT BLANK VALUE

BLANK VALUE = 21.33574 micrograms carbon

BLANK FACTOR = 21.33574 / 10.9978 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(377.3 --2772.271) (1)/(1) = +3.150E+03 g/L Carbon
(377.3 --2772.271) (1)/(1) (12) = +2.625E+02 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

Phase Recalculate

SS = .0698g + .100 ml ^{SPK} 26W/2 F 3642

LABCORE Completed Worklist Report for Worklist# 26753

Analyst: krm Instrument: CARB1 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@TICTOC1 TIC-02	SOLID	6.02E+02	6.15E+2	102.159 % Recovery	
1 STD	0	@TICTOC1 TOC-02	SOLID	3.00E+03	2.79E+3	93.000 % Recovery	
2 BLNK	0	@TICTOC1 TIC-02	SOLID	1	2.90E+0	2.900 ug/g	
2 BLNK	0	@TICTOC1 TOC-02	SOLID	1	8.50E+0	8.500 ug/g	
3 SAMPLE	S98T002912 0	@TICTOC1 TIC-02	SOLID	N/A	1.68E+03	5.000 ug/g	
3 SAMPLE	S98T002912 0	@TICTOC1 TOC-02	SOLID	N/A	4.12E+02	40.000 ug/g	
4 DUP	S98T002912 0	@TICTOC1 TIC-02	SOLID	1.68E+3	1.77E+3	5.217 RPD	
4 DUP	S98T002912 0	@TICTOC1 TOC-02	SOLID	4.12E+2	4.51E+2	9.038 RPD	
5 SPK	S98T002912 0	@TICTOC1 TIC-02	SOLID	1.00E+02	8.65E+01	86.500 % Recovery	
5 SPK	S98T002912 0	@TICTOC1 TOC-02	SOLID	1.00E+02	9.05E+01	90.500 % Recovery	
6 SAMPLE	S98T002913 0	@TICTOC1 TIC-02	SOLID	N/A	1.30E+03	5.000 ug/g	
6 SAMPLE	S98T002913 0	@TICTOC1 TOC-02	SOLID	N/A	6.65E+02	40.000 ug/g	
7 DUP	S98T002913 0	@TICTOC1 TIC-02	SOLID	1.30E+3	1.32E+3	1.527 RPD	
7 DUP	S98T002913 0	@TICTOC1 TOC-02	SOLID	6.65E+2	6.31E+2	5.247 RPD	
8 SAMPLE	S98T002914 0	@TICTOC1 TIC-02	SOLID	N/A	1.10E+04	5.000 ug/g	
8 SAMPLE	S98T002914 0	@TICTOC1 TOC-02	SOLID	N/A	8.59E+03	40.000 ug/g	
9 DUP	S98T002914 0	@TICTOC1 TIC-02	SOLID	1.10E+4	9.03E+3	19.670 RPD	
9 DUP	S98T002914 0	@TICTOC1 TOC-02	SOLID	8.59E+3	6.11E+3	33.741 RPD	
10 SPK	S98T002914 0	@TICTOC1 TIC-02	SOLID	1.00E+02	9.01E+01	90.100 % Recovery	
10 SPK	S98T002914 0	@TICTOC1 TOC-02	SOLID	1.00E+02	0.00E+00	0.000 % Recovery	

Final page for worklist# 26753

Analyst Signature _____ Date _____

[Signature] 11/24/98
Analyst Signature _____ Date _____

[Signature] 11/1/98
Reviewer Signature _____ Date _____

Sample S98T002914 will
be Rerun Due to SPK/DUP
failure. m/c

LABCORE Data Entry Template for Worklist# 26753

Analyst: KRM Instrument: CARB2 Book# 26753

Method: LA-342-100 Rev/Mod F-1

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	SOLID		
2 BLNK			@TICTOC1	SOLID		
3 SAMPLE	S98T002912 0		@TICTOC1	SOLID	98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002912 0		@TICTOC1	SOLID		
5 SPK	S98T002912 0		@TICTOC1	SOLID		
6 SAMPLE	S98T002913 0		@TICTOC1	SOLID	98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T002913 0		@TICTOC1	SOLID		
8 SAMPLE	S98T002914 0		@TICTOC1	SOLID	98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
9 DUP	S98T002914 0		@TICTOC1	SOLID		
10 SPK	S98T002914 0		@TICTOC1	SOLID		

Final page for worklist # 26753

Analyst Signature Date 11-15-98

Analyst Signature Date 11/24/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3644

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	624.5	582.8
26753	μg of Carbon from Baseline (C2)	9.9	24.4
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98005354			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	TIC	TOC
12:00 AM	QC Actual in $\mu\text{g}/\text{mL}$	5	40
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
BY-105	Percent Standard Recovery	6.15E+02	2.79E+03
		102.1	93.1

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	<i>NK</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3645

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	ug of Carbon in Sample (C1)	12.8	32.9
26753	ug of Carbon from Baseline (C2)	9.9	24.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005354			
Rerun	ug of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 AM			
Sample Point	ug of Carbon	2.90E+00	8.50E+00
BY-105			

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	<i>MA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3646

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0678	0.0678
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	123.9	52.3
26753	µg of Carbon from Baseline (C2)	9.9	24.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005354			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002912			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
		5	40
Sample Point	µg of Carbon/g	1.68E+03	4.12E+02
BY-105			

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	<i>N/A</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.2714	0.2714
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	489.1	146.7
26753	µg of Carbon from Baseline (C2)	9.9	24.4
Test Code	Known µg of C from Original Sample	1.68E+3	4.12E+2
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005354			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002912			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
Sample Point	µg of Carbon/g	1.77E+03	4.51E+02
BY-105			

Data Entered By:

CJO

Date:

11/24/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3648

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0678	0.0678
Work List	Final Coulometer Reading in μg (C1)	123.9	52.3
26753	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.2641	0.2641
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in μg (C2)	714.4	404.6
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in $\mu\text{g}/\text{mL}$ (SPK CONC)	602	3000
98005354	μg C in baseline (BL)	9.9	24.4
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = $((C2-BL) - (C1-BL)) * (SPK SS) / SS / ((SPK CONC) * (SPK VOL)) * 100$		
S98T002912			
Instrument Code	QC Actual in $\mu\text{g}/\text{mL}$ = Spike Value ($\mu\text{g}/\text{mL}$)		
CARB1	QC Found in $\mu\text{g}/\text{mL}$ = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time			
		TIC	TOC
	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.21E+02	2.72E+03
BY-105	Percent Spike Recovery	86.5	90.5

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.2836	0.2836
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	376.7	211.3
26753	µg of Carbon from Baseline (C2)	7.7	22.8
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005354			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002913			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
Sample Point	µg of Carbon/g	1.30E+03	6.65E+02
BY-105			

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3650

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.4453	0.4453
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	595.3	304
26753	µg of Carbon from Baseline (C2)	7.7	22.8
Test Code	Known µg of C from Original Sample	1.30E+3	6.65E+2
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005354			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002913			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
KRM			
Date Complete			
11/24/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
Sample Point	µg of Carbon/g	1.32E+03	6.31E+02
BY-105			

Data Entered By:	CJO	Date:	11/24/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3651

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE1

Date: 11/08/98

Time: 09:23:58

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE1

Blank Value = N/A

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.50	0.00
2	1.01	0.80	37.50
3	1.50	1.30	38.46
4	2.00	1.80	27.78
5	2.50	2.40	25.00
6	3.00	2.90	17.24
7	3.50	3.60	19.44
8	4.00	3.90	7.69
9	4.50	4.30	9.30
10	5.00	4.90	12.24
11	5.50	5.20	5.77
12	6.00	5.60	7.14
13	6.50	6.10	8.20
14	7.00	6.50	6.15
15	7.50	6.90	5.80
16	8.00	7.40	6.76
17	8.50	7.80	5.13
18	9.00	8.20	4.88
19	9.50	8.70	5.75
20	10.00	9.00	3.33
21	10.50	9.50	5.26
22	11.00	9.90	4.04

BLANK VALUE = 9.9 micrograms carbon

BLANK FACTOR = 9.9 / 10.99786 =

+9.0E-01

ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3612 TO 3613.

Sample Run By:

KR MONTEITH

00000

3652

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TIC/TOC REV 2:0

<<< BLANK ANALYSIS >>>

Sample: BASE1

Date: 11/08/98

Time: 18:36:22

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE1
 Blank Value = N/A

Analyst : KR MONTEITH
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.60	50.00
3	1.51	1.00	40.00
4	2.00	1.40	28.57
5	2.50	1.90	26.32
6	3.00	2.30	17.39
7	3.50	2.70	14.81
8	4.00	3.10	12.90
9	4.50	3.50	11.43
10	5.00	3.80	7.89
11	5.50	4.10	7.32
12	6.00	4.40	6.82
13	6.50	4.70	6.38
14	7.00	5.00	6.00
15	7.50	5.40	7.41
16	8.00	5.70	5.26
17	8.50	6.00	5.00
18	9.00	6.30	4.76
19	9.50	6.70	5.97
20	10.00	7.00	4.29
21	10.50	7.30	4.11
22	11.00	7.70	5.19

BLANK VALUE = 7.7 micrograms carbon

BLANK FACTOR = 7.7 / 10.99878 = +7.0E-01 ug/min Carbon

Sample Run By:

KR MONTEITH

00000

3653

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE1

Date: 11/08/98

Time: 09:38:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE1

Blank Value = N/A

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.90	78.95
3	1.51	6.50	70.77
4	2.00	11.30	42.48
5	2.50	14.30	20.98
6	3.00	15.80	9.49
7	3.50	16.90	6.51
8	4.00	17.50	3.43
9	4.50	18.30	4.37
10	5.00	18.70	2.14
11	5.50	19.40	3.61
12	6.00	19.90	2.51
13	6.50	20.30	1.97
14	7.00	21.00	3.33
15	7.50	21.30	1.41
16	8.00	21.80	2.29
17	8.50	22.20	1.80
18	9.00	22.60	1.77
19	9.50	23.10	2.16
20	10.00	23.60	2.12
21	10.50	24.10	2.07
22	11.00	24.40	1.23

BLANK VALUE = 24.4 micrograms carbon

BLANK FACTOR = 24.4 / 10.99799 = +2.22E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3654

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TIC/TOC REV 2300

<<< BLANK ANALYSIS >>>

Sample: BASE1

Date: 11/08/98

Time: 19:00:38

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE1
 Blank Value = N/A

Analyst : KR MONTEITH
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	0.40	0.00
2	1.00	0.80	50.00
3	1.50	3.30	75.76
4	2.00	8.40	60.71
5	2.50	12.30	31.71
6	3.00	14.30	13.99
7	3.50	15.40	7.14
8	4.00	16.40	6.10
9	4.50	17.10	4.09
10	5.00	17.80	3.93
11	5.50	18.50	3.78
12	6.00	19.20	3.65
13	6.50	19.70	2.54
14	7.00	20.00	1.50
15	7.50	20.40	1.96
16	8.00	20.80	1.92
17	8.50	21.10	1.42
18	9.00	21.50	1.86
19	9.50	21.80	1.38
20	10.00	22.10	1.36
21	10.50	22.40	1.34
22	11.00	22.80	1.75

BLANK VALUE = 22.8 micrograms carbon

BLANK FACTOR = 22.8 / 10.99866 = +2.07E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3655

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD1

Date: 11/08/98

Time: 09:54:48

Sample Size = 1000 uL

Analyst : KR MONTEITH

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .9 ug/minute C

% Difference = 10

== Reading == Analysis Time == Coulometer == % Difference ==

1	0.51	0.60	0.00
2	1.01	1.10	45.45
3	1.50	13.50	91.85
4	2.00	119.10	88.66
5	2.50	293.50	59.42
6	3.00	440.60	33.39
7	3.50	533.70	17.44
8	4.00	580.80	8.11
9	4.50	603.20	3.71
10	5.00	612.20	1.47
11	5.50	616.10	0.63
12	6.00	617.80	0.28
13	6.50	619.00	0.19
14	7.00	619.90	0.15
15	7.50	620.60	0.11
16	8.00	621.30	0.11
17	8.50	621.70	0.06
18	9.00	622.40	0.11
19	9.50	622.90	0.08
20	10.00	623.40	0.08
21	10.50	624.00	0.10
22	11.00	624.50	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 9.898187 micrograms carbon

BLANK FACTOR = 9.898187 / 10.99799 = +9.0E-01 ug/min Carbon

SAMPLE RESULTS:

(624.5 - 9.898077)(1)/(1000) = +6.146E-01 g/L Carbon

(624.5 - 9.898077)(1)/(1000)(12) = +5.122E-02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3656

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD1

Date: 11/08/98

Time: 19:15:01

Sample Size = 1000 uL   Analyst :   KR MONTEITH
Dil Factor = 1   Min Readings = 22
Blank ID # =   Max Readings = 22
Blank Value = .7 ug/minute C   % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	2.00	75.00
3	1.51	67.70	97.05
4	2.00	227.20	70.20
5	2.50	392.40	42.10
6	3.00	503.80	22.11
7	3.50	562.60	10.45
8	4.00	588.20	4.35
9	4.50	598.40	1.70
10	5.00	602.30	0.65
11	5.50	603.80	0.25
12	6.00	604.80	0.17
13	6.50	605.40	0.10
14	7.00	606.00	0.10
15	7.50	606.50	0.08
16	8.00	606.90	0.07
17	8.50	607.40	0.08
18	9.00	607.80	0.07
19	9.50	608.20	0.07
20	10.00	608.60	0.07
21	10.50	608.90	0.05
22	11.00	609.30	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 7.69906 micrograms carbon

BLANK FACTOR = 7.69906 / 10.99866 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(609.3 - 7.699145 | (1)/(1000) = +6.016E-01 g/L Carbon
(609.3 - 7.699145 | (1)/(1000) (12) = +5.013E-02 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3657

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD1

Date: 11/08/98

Time: 10:37:27

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.22 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	8.80	93.18
3	1.51	139.90	93.71
4	2.00	395.30	64.61
5	2.50	519.10	23.85
6	3.00	555.50	6.55
7	3.50	566.60	1.96
8	4.00	570.90	0.75
9	4.50	572.90	0.35
10	5.00	574.50	0.28
11	5.50	575.50	0.17
12	5.00	576.30	0.14
13	5.50	577.30	0.17
14	7.00	577.90	0.10
15	7.50	578.70	0.14
16	8.00	579.40	0.12
17	8.50	580.10	0.12
18	9.00	580.60	0.09
19	9.50	581.20	0.10
20	10.00	581.80	0.10
21	10.50	582.30	0.09
22	11.00	582.80	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 24.41553 micrograms carbon

BLANK FACTOR = 24.41553 / 10.99799 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(582.8 - 24.41729) (1) / (200) = +2.792E+00 g/L Carbon
 (582.8 - 24.41729) (1) / (200) (12) = +2.327E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3658

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD1

Date: 11/08/98

Time: 19:33:16

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.07 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.70	0.00
2	1.01	1.10	36.36
3	1.51	3.00	63.33
4	2.00	13.20	77.27
5	2.50	84.40	84.36
6	3.00	330.50	74.46
7	3.50	506.20	34.71
8	4.00	555.00	8.79
9	4.50	569.50	2.55
10	5.00	574.40	0.85
11	5.50	576.70	0.40
12	6.00	578.00	0.22
13	6.50	579.10	0.19
14	7.00	579.90	0.14
15	7.50	580.70	0.14
16	8.00	581.40	0.12
17	8.50	582.00	0.10
18	9.00	582.60	0.10
19	9.50	583.20	0.10
20	10.00	583.70	0.09
21	10.50	584.20	0.09
22	11.00	584.70	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 22.76722 micrograms carbon

BLANK FACTOR = 22.76722 / 10.99866 = +2.1E+00 ug/min Carbon

SAMPLE RESULTS:

(584.7 - 22.7657) (1)/(200) = +2.810E+00 g/L Carbon
 (584.7 - 22.7657) (1)/(200) (12) = +2.341E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3659

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 12.0

Sample: BLK1

Date: 11/08/98

Time: 10:50:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .9 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.50	60.00
3	1.51	2.30	54.55
4	2.00	4.70	29.79
5	2.50	5.30	11.32
6	3.00	6.00	11.67
7	3.50	6.40	6.25
8	4.00	6.80	5.88
9	4.50	7.30	6.85
10	5.00	7.80	6.41
11	5.50	8.10	3.70
12	6.00	8.50	4.71
13	6.50	8.90	4.49
14	7.00	9.50	6.32
15	7.50	9.80	3.06
16	8.00	10.30	4.85
17	8.50	10.70	3.74
18	9.00	11.10	3.60
19	9.50	11.60	4.31
20	10.00	12.00	3.33
21	10.50	12.50	4.00
22	11.00	12.80	2.34

USER INPUT BLANK VALUE

BLANK VALUE = 9.898187 micrograms carbon

BLANK FACTOR = 9.898187 / 10.99799 = +9.0E-01 ug/min Carbon

SAMPLE RESULTS:

(12.8 - 9.898187) (1)/(1) =

+2.90E+00 g/L Carbon

(12.8 - 9.898187) (1)/(1) (12) =

+2.42E-01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3660

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK1

Date: 11/08/98

Time: 19:45:33

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	1.80	83.33
3	1.50	3.60	50.00
4	2.00	4.50	20.00
5	2.50	5.00	10.00
6	3.00	5.50	9.09
7	3.50	5.90	6.78
8	4.00	6.20	4.84
9	4.50	6.60	6.06
10	5.00	6.90	4.35
11	5.50	7.30	5.48
12	6.00	7.70	5.19
13	6.50	8.00	3.75
14	7.00	8.40	4.76
15	7.50	8.70	3.45
16	8.00	9.00	3.33
17	8.50	9.30	3.23
18	9.00	9.70	4.12
19	9.50	10.00	3.00
20	10.00	10.30	2.91
21	10.50	10.70	3.74
22	11.00	11.00	2.73

USER INPUT BLANK VALUE

BLANK VALUE = 7.69906 micrograms carbon

BLANK FACTOR = 7.69906 / 10.99866 =

+7.0E-01

ug/min Carbon

SAMPLE RESULTS:

(11 - 7.699145) (1)/(1) =

+3.30E+00

g/L Carbon

(11 - 7.699145) (1)/(1) (12) =

+2.75E-01

Molar Carbon

Sample Run By:

KR MONTEITH

00000

3661

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK1

Date: 11/08/98

Time: 11:03:10

Sample Size = 1 uL

Analyst : KR MONTEITH

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.22 ug/minute C

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.04	4.40	88.64
3	1.54	10.90	59.63
4	2.04	16.90	35.50
5	2.54	21.10	19.91
6	3.04	23.30	9.44
7	3.54	24.80	6.05
8	4.04	25.80	3.88
9	4.54	26.50	2.64
10	5.04	27.00	1.85
11	5.54	27.60	2.17
12	6.04	28.20	2.13
13	6.54	28.60	1.40
14	7.04	29.20	2.05
15	7.54	29.60	1.35
16	8.04	30.20	1.99
17	8.54	30.60	1.31
18	9.04	31.10	1.61
19	9.54	31.50	1.27
20	10.04	32.00	1.56
21	10.54	32.50	1.54
22	11.04	32.90	1.22

USER INPUT BLANK VALUE

BLANK VALUE = 24.41553 micrograms carbon

BLANK FACTOR = 24.41553 / 10.99799 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(32.9 - 24.49886) (1)/(1) = +8.40E+00 g/L Carbon
 (32.9 - 24.49886) (1)/(1)(12) = +7.00E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3662

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK1

Date: 11/08/98

Time: 20:07:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.07 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.90	73.68
3	1.50	4.40	56.82
4	2.00	9.60	54.17
5	2.50	15.70	38.85
6	3.00	21.00	25.24
7	3.50	24.30	13.58
8	4.00	26.30	7.60
9	4.50	27.60	4.71
10	5.00	28.90	4.50
11	5.50	30.40	4.93
12	6.00	31.50	3.49
13	6.50	32.80	3.96
14	7.00	34.00	3.53
15	7.50	35.40	3.95
16	8.00	36.40	2.75
17	8.50	37.40	2.67
18	9.00	38.10	1.84
19	9.50	39.00	2.31
20	10.00	39.70	1.76
21	10.50	40.30	1.49
22	11.00	40.80	1.23

USER INPUT BLANK VALUE

BLANK VALUE = 22.76722 micrograms carbon

BLANK FACTOR = 22.76722 / 10.99866 = +2.1E+00 ug/min Carbon

SAMPLE RESULTS:

(40.8 - 22.76545) (1)/(1) = +1.80E+01 g/L Carbon
 (40.8 - 22.76545) (1)/(1) (12) = +1.50E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3663

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912

Date: 11/08/98

Time: 15:09:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .9 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	3.50	0.00
2	1.01	4.50	22.22
3	1.50	15.80	71.52
4	2.00	47.20	66.53
5	2.50	77.60	39.18
6	3.00	98.30	21.06
7	3.50	109.30	10.06
8	4.00	113.90	4.04
9	4.50	115.90	1.73
10	5.00	117.10	1.02
11	5.50	117.90	0.68
12	6.00	118.50	0.51
13	6.50	119.20	0.59
14	7.00	119.60	0.33
15	7.50	120.20	0.50
16	8.00	120.70	0.41
17	8.50	121.20	0.41
18	9.00	121.70	0.41
19	9.50	122.40	0.57
20	10.00	123.00	0.49
21	10.50	123.50	0.40
22	11.00	123.90	0.32

USER INPUT BLANK VALUE

BLANK VALUE = 9.891187 micrograms carbon

BLANK FACTOR = 9.81187 / 10.99799 = +9.0E-01 ug/min Carbon

SAMPLE RESULTS:

{ 123.9 - 9.898956 } (1)/(1) = +1.140E+02 g/L Carbon
{ 123.9 - 9.898956 } (1)/(1)(12) = +9.500E+00 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3664

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912

Date: 11/08/98

Time: 15:25:42

Sample Size = 1 uL

Analyst : KR MONTEITH

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.22 ug/minute C

% Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference
1	0.51		0.50	0.00
2	1.01		5.40	90.74
3	1.51		21.00	74.29
4	2.01		32.00	34.38
5	2.50		37.20	13.98
6	3.00		39.80	6.53
7	3.50		41.60	4.33
8	4.00		42.90	3.03
9	4.50		44.20	2.94
10	5.00		45.20	2.21
11	5.50		45.80	1.31
12	6.00		46.80	2.14
13	6.50		47.30	1.06
14	7.00		47.90	1.25
15	7.50		48.60	1.44
16	8.00		49.10	1.02
17	8.50		49.70	1.21
18	9.00		50.20	1.00
19	9.50		50.80	1.18
20	10.00		51.30	0.97
21	10.50		51.80	0.97
22	11.00		52.30	0.96

USER INPUT BLANK VALUE

BLANK VALUE = 24.41553 micrograms carbon

BLANK FACTOR = 24.41553 / 10.99799 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(52.3 - 24.41743) (1)/(1) = +2.79E+01 g/L Carbon
 (52.3 - 24.41743) (1)/(1) (12) = +2.32E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3665

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912DUP

Date: 11/08/98

Time: 15:38:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .9 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	7.40	93.24
3	1.51	86.00	91.40
4	2.01	217.50	60.46
5	2.50	331.40	34.37
6	3.00	406.50	18.49
7	3.50	448.80	9.40
8	4.00	468.00	4.10
9	4.50	475.80	1.64
10	5.00	479.10	0.69
11	5.50	480.60	0.31
12	6.00	481.70	0.23
13	6.50	482.70	0.21
14	7.00	483.50	0.17
15	7.50	484.20	0.14
16	8.00	485.20	0.21
17	8.50	485.80	0.12
18	9.00	486.30	0.10
19	9.50	487.10	0.16
20	10.00	487.60	0.10
21	10.50	488.30	0.14
22	11.00	489.10	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 9.891187 micrograms carbon

BLANK FACTOR = 9.811817 / 10.99799 = +9.0E-01 ug/min Carbon

SAMPLE RESULTS:

(489.1 - 9.898956) (1)/(1) = +4.792E+02 g/L Carbon
 (489.1 - 9.898956) (1)/(1) (12) = +3.993E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3666

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912DUP

Date: 11/08/98

Time: 15:51:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.22 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.90	0.00
2	1.01	13.40	93.28
3	1.51	71.70	81.31
4	2.01	108.80	34.10
5	2.50	122.60	11.26
6	3.00	129.00	4.96
7	3.50	132.30	2.49
8	4.00	134.70	1.78
9	4.50	136.70	1.46
10	5.00	138.30	1.16
11	5.50	139.30	0.72
12	6.00	140.20	0.64
13	6.50	141.00	0.57
14	7.00	141.70	0.49
15	7.50	142.50	0.56
16	8.00	143.10	0.42
17	8.50	143.90	0.56
18	9.00	144.40	0.35
19	9.50	145.00	0.41
20	10.00	145.70	0.48
21	10.50	146.20	0.34
22	11.00	146.70	0.34

USER INPUT BLANK VALUE

BLANK VALUE = 24.41553 micrograms carbon

BLANK FACTOR = 24.41553 / 10.99799 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(146.7 - 24.41973) (1) / (1) = +1.223E+02 g/L Carbon
 (146.7 - 24.41973) (1) / (1) (12) = +1.019E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3667

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912SPK

Date: 11/08/98

Time: 16:04:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .9 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	2.00	75.00
3	1.50	60.00	96.67
4	2.00	230.60	73.98
5	2.50	421.60	45.30
6	3.00	558.40	24.50
7	3.50	638.10	12.49
8	4.00	678.30	5.93
9	4.50	695.30	2.44
10	5.00	701.80	0.93
11	5.50	704.70	0.41
12	6.00	706.20	0.21
13	6.50	707.40	0.17
14	7.00	708.30	0.13
15	7.50	709.20	0.13
16	8.00	710.00	0.11
17	8.50	710.80	0.11
18	9.00	711.50	0.10
19	9.50	712.10	0.08
20	10.00	712.80	0.10
21	10.50	713.60	0.11
22	11.00	714.40	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 9.898187 micrograms carbon

BLANK FACTOR = 9.88187 / 10.99799 = +9.0E-01 ug/min Carbon

SAMPLE RESULTS:

(714.4 - 9.898956) (1)/(1) = +7.045E+02 g/L Carbon
 (714.4 - 9.898956) (1)/(1) (12) = +5.871E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3668

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2912SPK

Date: 11/08/98

Time: 16:16:49

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.22 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	21.10	97.63
3	1.50	153.40	86.25
4	2.00	306.60	49.97
5	2.50	361.90	15.28
6	3.00	379.30	4.59
7	3.50	386.00	1.74
8	4.00	389.50	0.90
9	4.50	392.00	0.64
10	5.00	394.10	0.53
11	5.50	395.40	0.33
12	6.00	396.60	0.30
13	6.50	397.80	0.30
14	7.00	399.00	0.30
15	7.50	399.70	0.18
16	8.00	400.60	0.22
17	8.50	401.20	0.15
18	9.00	401.90	0.17
19	9.50	402.70	0.20
20	10.00	403.20	0.12
21	10.50	404.00	0.20
22	11.00	404.60	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 24.41553 micrograms carbon

BLANK FACTOR = 24.41553 / 10.99799 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(404.6 - 24.41743) (1)/(1) = +3.802E+02 g/L Carbon
 (404.6 - 24.41743) (1)/(1)(12) = +3.168E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3669

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

Sample: S98T2711-1

Date: 11/08/98

Time: 20:22:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	10.70	42.86
3	1.50	2.80	75.00
4	2.00	28.90	90.31
5	2.50	100.50	71.24
6	3.00	196.60	48.88
7	3.50	275.00	28.51
8	4.00	322.70	14.78
9	4.50	348.50	7.40
10	5.00	360.90	3.44
11	5.50	366.80	1.61
12	6.00	369.60	0.76
13	6.50	371.10	0.40
14	7.00	372.10	0.27
15	7.50	372.90	0.21
16	8.00	373.50	0.16
17	8.50	374.20	0.19
18	9.00	374.70	0.13
19	9.50	375.30	0.16
20	10.00	375.70	0.11
21	10.50	376.30	0.16
22	11.00	376.70	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 7.69806 micrograms carbon

BLANK FACTOR = 7.69806 / 10.99866 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(376.7 - 7.698547) (1)/(1) = +3.690E+02 g/L Carbon
 (376.7 - 7.698547) (1)/(1) (12) = +3.075E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3670

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T2913-1

Date: 11/08/98

Time: 20:34:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.07 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.50	6.20	85.48
4	2.00	46.60	86.70
5	2.50	116.60	60.03
6	3.00	161.30	27.71
7	3.50	180.70	10.74
8	4.00	189.40	4.59
9	4.50	194.40	2.57
10	5.00	198.10	1.87
11	5.50	200.70	1.30
12	6.00	202.80	1.04
13	6.50	204.40	0.78
14	7.00	205.70	0.63
15	7.50	206.80	0.53
16	8.00	207.80	0.48
17	8.50	208.60	0.38
18	9.00	209.40	0.38
19	9.50	210.10	0.33
20	10.00	210.40	0.14
21	10.50	210.90	0.24
22	11.00	211.30	0.19

USER INPUT BLANK VALUE

BLANK VALUE = 22.76722 micrograms carbon

BLANK FACTOR = 22.76722 / 10.99866 = +2.1E+00 ug/min Carbon

SAMPLE RESULTS:

(211.3 - 22.76747) (1)/(1) = +1.885E+02 g/L Carbon
 (211.3 - 22.76747) (1)/(1)(12) = +1.571E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3671

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T2913DUP-1

Date: 11/08/98

Time: 20:47:32

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst :

KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.70	0.00
2	1.01	1.00	30.00
3	1.50	1.70	41.18
4	2.00	9.90	82.83
5	2.50	52.60	81.18
6	3.00	150.50	65.05
7	3.50	295.50	49.07
8	4.00	418.00	29.31
9	4.50	495.90	15.71
10	5.00	542.20	8.54
11	5.50	567.60	4.47
12	6.00	579.90	2.12
13	6.50	585.60	0.97
14	7.00	588.20	0.44
15	7.50	590.00	0.31
16	8.00	591.20	0.20
17	8.50	592.00	0.14
18	9.00	592.90	0.15
19	9.50	593.40	0.08
20	10.00	594.10	0.12
21	10.50	594.70	0.10
22	11.00	595.30	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 7.6906 micrograms carbon

BLANK FACTOR = 7.6906 / 10.99866 =

+7.0E-01

ug/min Carbon

SAMPLE RESULTS:

(595.3 - 7.699145 (1)/(1) =

+5.876E+02 g/L Carbon

(595.3 - 7.699145 (1)/(1) (12) =

+4.897E+01 Molar Carbon

Sample Run By:

KR MONTEITH

00000

3672

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T2913DUP-1

Date: 11/08/98

Time: 21:09:39

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.07 ug/minute C

Analyst : KR MONTEITH

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	3.00	83.33
3	1.51	46.40	93.53
4	2.00	130.00	64.31
5	2.50	214.30	39.34
6	3.00	254.50	15.80
7	3.50	272.10	6.47
8	4.04	281.00	3.17
9	4.54	286.70	1.99
10	5.04	290.40	1.27
11	5.54	293.40	1.02
12	6.04	295.30	0.64
13	6.53	297.00	0.57
14	7.03	298.50	0.50
15	7.53	299.60	0.37
16	8.03	300.50	0.30
17	8.53	301.40	0.30
18	9.03	302.00	0.20
19	9.53	302.70	0.23
20	10.03	303.10	0.13
21	10.53	303.60	0.16
22	11.03	304.00	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 22.75722 micrograms carbon

BLANK FACTOR = 22.75722 / 10.99866 = +2.1E+00 ug/min Carbon

SAMPLE RESULTS:

(304 - 22.83393) (1) / (1) = +2.812E+02 g/L Carbon
 (304 - 22.83393) (1) / (1) (12) = +2.343E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

KR MONTEITH

00000

3673

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 12 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 3674 THRU 3965)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	TOTAL INORGANIC CARBON/TOTAL ORGANIC CARBON ANALYSIS, TIC/TOC WORKLIST 26755 THRU 28635		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635
Specific Gravity Analysis (SpG)	
SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684
Hydroxide Analysis (OH)	
OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775
Ammonia Analysis (NH3)	
NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992
Chromium VI Analysis (CRVI)	
CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

Sec 3
of 17

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
Continued

See 4
of 17

See 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26755

Analyst: jds

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	6.02E+02	6.13E+2	101.827 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.89E+3	96.333 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	6.10E+0	6.100 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	1.55E+1	15.500 ug/g	
3 SAMPLE	S98T002982	0	@TICTOC1 TIC-02	SOLID	N/A	6.41E+03	5.000 ug/g	
3 SAMPLE	S98T002982	0	@TICTOC1 TOC-02	SOLID	N/A	3.72E+02	40.000 ug/g	
4 DUP	S98T002982	0	@TICTOC1 TIC-02	SOLID	6.41E+3	5.59E+3	13.667 RPD	
4 DUP	S98T002982	0	@TICTOC1 TOC-02	SOLID	3.72E+2	3.08E+2	18.824 RPD	
5 SAMPLE	S98T002983	0	@TICTOC1 TIC-02	SOLID	N/A	2.95E+03	5.000 ug/g	
5 SAMPLE	S98T002983	0	@TICTOC1 TOC-02	SOLID	N/A	1.31E+03	40.000 ug/g	
6 DUP	S98T002983	0	@TICTOC1 TIC-02	SOLID	2.95E+3	3.42E+3	14.757 RPD	
6 DUP	S98T002983	0	@TICTOC1 TOC-02	SOLID	1.31E+3	1.77E+3	29.870 RPD	

Final page for worklist# 26755

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

Sample S98T002983
out for Rerun Dup failure
mjc

10/23/98 09:39
A-0004-1

LABCORE Data Entry Template for Worklist# 26755

Analyst: Sds Instrument: CARB2 Book# 25N/2D
25N/12FMethod: LA-342-100 Rev/Mod F-1

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@TICTOC1	SOLID		
2	BLNK				@TICTOC1	SOLID		
3	SAMPLE	S98T002982	0		@TICTOC1	SOLID	98000509	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
4	DUP	S98T002982	0		@TICTOC1	SOLID		
5	SAMPLE	S98T002983	0		@TICTOC1	SOLID	98000509	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
6	DUP	S98T002983	0		@TICTOC1	SOLID		

Final page for worklist # 26755

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	623	611.1
26755	µg of Carbon from Baseline (C2)	9.9	34.1
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000

Matrix
LIQUIDBatch Number
98005356Rerun
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
CJOChemist
MJLAnalyst
JISDate Complete
11/16/98Analysis Date
11/13/98Analysis Time
05:00 AMSample Point
BY-105 (2)

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

Method Detection Limit in µg/mL

QC Actual in µg/mL

QC Found in µg/mL

Percent Standard Recovery

TIC	TOC
5	40
6.02E+02	3.00E+03
6.13E+02	2.89E+03
101.8	96.2

Data Entered By:

Signature of Chemist:

CJO

Date: 11/16/98

Date:

STANDARD.WB1 REV 1.0

342100ML

3676

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	3.8	18.6
26755	µg of Carbon from Baseline (C2)	9.9	34.1
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005356			
Rerun	µg of Carbon = [C1-C2]		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
JIS			
Date Complete			
11/16/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 AM			
Sample Point	µg of Carbon	6.10E+00	1.55E+01
BY-105 (2)			

Data Entered By:

CJO

Date: 11/16/98

Signature of Chemist:

Date:

BLANK.WB1 REV 1.0

342100ML

3677

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.1217	0.1217
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	790.1	79.4
26755	µg of Carbon from Baseline (C2)	9.9	34.1
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005356			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002982			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
11/16/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 AM			
Sample Point	µg of Carbon/g	6.41E+03	3.72E+02
BY-105 (2)			

Data Entered By:	CJO	Date:	11/16/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.1155	0.1155
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	655.8	69.7
26755	µg of Carbon from Baseline (C2)	9.9	34.1
Test Code	Known µg of C from Original Sample	6.41E+3	3.72E+2
@TIC/TOC1			
Matrix	µg of Carbon/g = (C1-C2) * DF / SS µg of Carbon/g for TIC = 5 if C1 < C2 µg of Carbon/g for TOC = 40 if C1 < C2		
SOLID			
Batch Number			
98005356			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T002982			
Instrument Code			
CARB2			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
JIS			
Date Complete			
11/16/98			
Analysis Date			
11/13/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 AM			
Sample Point	µg of Carbon/g	5.59E+03	3.08E+02
BY-105 (2)			

Data Entered By:	CJO	Date:	11/16/98
Signature of Chemist:	<i>NA</i>	Date:	

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TIC/OC REV 2.0
<<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/12/88

Time: 19:46:53

Sample Size = 1 uL
Dil Factor = 1
Blank ID # = BASE
Blank Value = N/AAnalyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	1.10	45.45
4	1.51	1.90	42.11
5	2.00	2.50	24.00
6	2.50	3.20	21.88
7	3.00	3.50	8.57
8	3.50	3.70	5.41
9	4.00	4.00	7.50
10	4.50	4.30	6.98
11	5.00	4.80	10.42
12	5.50	5.20	7.69
13	6.00	5.70	8.77
14	6.50	6.10	6.56
15	7.00	6.50	6.15
16	7.50	7.00	7.14
17	8.00	7.50	6.67
18	8.50	7.90	5.06
19	9.00	8.30	4.82
20	9.50	8.80	5.68
21	10.00	9.20	4.35
22	10.50	9.90	7.07

BLANK VALUE = 9.9 micrograms carbon
BLANK FACTOR = 9.9 / 10.49805 =

+9.4E-01 ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3680 TO 3689Sample Run By:

JD SPELLMAN

22222

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/12/98

Time: 20:06:04

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : JD SPELLMAN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	Analysis	Time	Coulometer	% Difference ==
1	0.08		0.00	0.00
2	0.51		2.00	100.00
3	1.01		5.40	68.75
4	1.51		12.30	47.97
5	2.01		18.00	31.67
6	2.50		21.50	16.28
7	3.00		23.80	9.66
8	3.50		25.30	5.93
9	4.00		26.20	3.44
10	4.50		27.10	3.32
11	5.00		27.70	2.17
12	5.50		28.40	2.46
13	6.00		29.10	2.41
14	6.50		29.80	2.35
15	7.00		30.30	1.65
16	7.50		30.80	1.62
17	8.00		31.40	1.91
18	8.50		32.10	2.18
19	9.00		32.50	1.23
20	9.50		33.00	1.52
21	10.00		33.50	1.49
22	10.50		34.10	1.76

BLANK VALUE = 34.1 micrograms carbon

BLANK FACTOR = 34.1 / 10.49915 = +3.25E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLEMAN

22222

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 12.0

Sample: STD

Date: 11/12/98

Time: 20:26:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .94 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.50	0.60	100.00
3	1.00	6.70	91.04
4	1.50	128.80	94.80
5	2.00	394.90	57.76
6	2.50	442.00	31.02
7	3.00	528.40	16.35
8	3.50	575.60	8.20
9	4.00	598.20	3.78
10	4.50	608.40	1.68
11	5.00	613.30	0.80
12	5.50	615.60	0.37
13	6.00	617.10	0.24
14	6.50	618.00	0.15
15	7.00	618.90	0.15
16	7.50	619.70	0.13
17	8.00	620.20	0.08
18	8.50	620.80	0.10
19	9.00	621.50	0.11
20	9.50	622.00	0.08
21	10.00	622.80	0.13
22	10.50	623.00	0.03

USER INPUT BLANK VALUE

BLANK VALUE = 9.869197 micrograms carbon

BLANK FACTOR = 9.869197 / 10.49915 = +9.4E-01 ug/min Carbon

SAMPLE RESULTS:

(623 - 9.868279) (1)/(1) =

+6.131E+02 g/L Carbon

(623 - 9.868279) (1)/(1)(12) =

+5.109E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 11/12/98

Time: 20:55:23

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.25 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	3.90	100.00
3	1.01	35.60	89.04
4	1.50	127.10	71.99
5	2.00	374.90	66.10
6	2.50	525.60	28.67
7	3.00	570.70	7.90
8	3.50	590.30	3.32
9	4.00	595.40	0.86
10	4.50	596.80	0.23
11	5.00	597.80	0.17
12	5.50	599.00	0.20
13	6.00	600.60	0.27
14	6.50	602.90	0.38
15	7.00	604.20	0.22
16	7.50	605.30	0.18
17	8.00	606.60	0.21
18	8.50	607.50	0.15
19	9.03	608.50	0.16
20	9.53	609.60	0.18
21	10.03	610.30	0.11
22	10.53	611.10	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 34.12222 micrograms carbon

BLANK FACTOR = 34.12222 / 10.49915 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(611.1 - 34.23132) (1) / (1) = -5.769E+02 g/L Carbon
 (611.1 - 34.23132) (1) / (1) (12) = -4.807E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/12/98

Time: 21:11:57

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .94 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	0.40	50.00
4	1.51	0.50	20.00
5	2.00	0.80	37.50
6	2.50	0.90	11.11
7	3.00	1.20	25.00
8	3.50	1.20	0.00
9	4.00	1.40	14.29
10	4.50	1.70	17.65
11	5.00	1.80	5.56
12	5.50	2.00	10.00
13	6.00	2.20	9.09
14	6.50	2.40	8.33
15	7.00	2.50	4.00
16	7.50	2.70	7.41
17	8.00	2.90	6.90
18	8.50	3.10	6.45
19	9.00	3.30	6.06
20	9.50	3.40	2.94
21	10.00	3.60	5.56
22	10.50	3.80	5.26

USER INPUT BLANK VALUE

BLANK VALUE = 9.869197 micrograms carbon

BLANK FACTOR = 9.869197 / 10.49915 = +9.4E-01 ug/min Carbon

SAMPLE RESULTS:

(3.8 - 9.867361) (1)/(1) =

< 5.00 E-3 g/L Carbon

(3.8 - 9.867361) (1)/(1) (12) =

< 4.17 E-4 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/12/98

Time: 21:30:34

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.25 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.50	1.30	100.00
3	1.00	3.90	66.67
4	1.50	7.00	44.29
5	2.00	9.90	29.29
6	2.50	11.70	15.38
7	3.00	12.90	9.30
8	3.50	13.40	3.73
9	4.00	14.00	4.29
10	4.50	14.50	3.45
11	5.00	14.90	2.68
12	5.50	15.30	2.61
13	6.00	15.70	2.55
14	6.50	16.00	1.88
15	7.00	16.40	2.44
16	7.50	16.70	1.80
17	8.00	17.00	1.76
18	8.50	17.50	2.86
19	9.00	17.70	1.13
20	9.50	18.00	1.67
21	10.00	18.30	1.64
22	10.50	18.60	1.61

USER INPUT BLANK VALUE

BLANK VALUE = 34.12222 micrograms carbon

BLANK FACTOR = 34.12222 / 10.49915 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(18.6 - 34.11588) (1)/(1) = < 5.00 E-3 g/L Carbon
 (18.6 - 34.11588) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002982

Date: 11/12/98

Time: 21:48:01

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = 1

Blank Value = .94 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	1.60	81.25
4	1.51	105.40	98.48
5	2.00	340.40	69.04
6	2.50	530.30	35.81
7	3.00	645.30	17.82
8	3.50	712.10	9.38
9	4.00	748.90	4.91
10	4.50	767.90	2.47
11	5.00	777.00	1.17
12	5.50	780.90	0.50
13	6.00	783.10	0.28
14	6.50	784.60	0.19
15	7.00	785.70	0.14
16	7.50	786.60	0.11
17	8.00	787.20	0.08
18	8.50	788.00	0.10
19	9.00	788.60	0.08
20	9.50	789.10	0.06
21	10.00	789.70	0.08
22	10.50	790.10	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 9.869197 micrograms carbon

BLANK FACTOR = 9.869197 / 10.49915 = +9.4E-01 ug/min Carbon

SAMPLE RESULTS:

(790.1 - 9.868279) (1)/(1) =

+7.802E+02 g/L Carbon

(790.1 - 9.868279) (1)/(1) (12) =

+6.502E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T002982

Date: 11/12/98

Time: 22:01:47

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.25 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	4.70	85.11
4	1.50	13.40	64.93
5	2.00	35.40	62.15
6	2.50	50.60	30.04
7	3.00	58.20	13.06
8	3.50	62.90	7.47
9	4.00	66.00	4.70
10	4.50	68.40	3.51
11	5.00	70.10	2.43
12	5.50	71.60	2.09
13	6.00	72.90	1.78
14	6.50	73.90	1.35
15	7.00	74.70	1.07
16	7.50	75.00	0.40
17	8.00	75.90	1.19
18	8.50	76.40	0.65
19	9.00	77.20	1.04
20	9.50	78.40	1.53
21	10.00	78.90	0.63
22	10.50	79.40	0.63

USER INPUT BLANK VALUE

BLANK VALUE = 34.12222 micrograms carbon

BLANK FACTOR = 34.12222 / 10.49915 =

+3.3E+00

ug/min Carbon

SAMPLE RESULTS:

(79.4 - 34.11865) (1)/(1) =

+4.53E+01

g/L Carbon

(79.4 - 34.11865) (1)/(1) (12) =

+3.77E+00

Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002982DUP

Date: 11/12/98

Time: 22:16:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .94 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	24.60	99.19
4	1.50	176.40	86.05
5	2.00	347.50	49.24
6	2.50	478.00	27.30
7	3.00	557.40	14.24
8	3.50	601.40	7.32
9	4.00	626.40	3.99
10	4.50	638.40	1.88
11	5.00	644.30	0.92
12	5.50	647.20	0.45
13	6.00	649.10	0.29
14	6.50	650.10	0.15
15	7.00	651.20	0.17
16	7.50	652.10	0.14
17	8.00	652.80	0.11
18	8.50	653.50	0.11
19	9.00	654.10	0.09
20	9.50	654.70	0.09
21	10.00	655.20	0.08
22	10.50	655.80	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 9.869197 micrograms carbon

BLANK FACTOR = 9.869197 / 10.49915 =

+9.4E-01

ug/min Carbon

SAMPLE RESULTS:

(655.8 - 9.869885) (1)/(1) =

+6.459E+02 g/L Carbon

(655.8 - 9.869885) (1)/(1) (12) =

+5.383E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002982

Date: 11/12/98

Time: 22:55:42

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.25 ug/minute C

Analyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.90	100.00
3	1.01	21.30	91.08
4	1.51	39.90	46.62
5	2.00	48.50	17.73
6	2.50	52.40	7.44
7	3.00	58.40	10.27
8	3.50	63.70	8.32
9	4.00	66.50	4.21
10	4.50	66.80	0.45
11	5.00	67.00	0.30
12	5.50	67.20	0.30
13	6.00	67.40	0.30
14	6.50	67.60	0.30
15	7.00	67.80	0.29
16	7.50	68.00	0.29
17	8.00	68.30	0.44
18	8.50	68.50	0.29
19	9.00	68.70	0.29
20	9.50	69.00	0.43
21	10.00	69.20	0.29
22	10.50	69.70	0.72

USER INPUT BLANK VALUE

BLANK VALUE = 34.12222 micrograms carbon

BLANK FACTOR = 34.12222 / 10.49915 = +3.3E+00 ug/min Carbon

SAMPLE RESULTS:

(69.7 - 34.12183) (1)/(1) = +3.56E+01 g/L Carbon
 (69.7 - 34.12183) (1)/(1) (12) = +2.96E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

LABCORE Completed Worklist Report for Worklist# 26846

Analyst: pjm

Instrument: CARB1

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0		@TICTOC1 TIC-02	LIQUID	1	1.30E+0	1.300	ug/mL
1 BLNK	0		@TICTOC1 TOC-02	LIQUID	1	1.70E+0	1.700	ug/mL
2 STD	0		@TICTOC1 TIC-02	LIQUID	6.02E+02	6.04E+2	100.332 % Recovery	
2 STD	0		@TICTOC1 TOC-02	LIQUID	3.00E+03	2.83E+3	94.333 % Recovery	
3 SAMPLE	S98T003010	0	@TICTOC1 TIC-02	LIQUID	N/A	4.83E+03	5.000	ug/mL
3 SAMPLE	S98T003010	0	@TICTOC1 TOC-02	LIQUID	N/A	2.38E+03	40.000	ug/mL
4 DUP	S98T003010	0	@TICTOC1 TIC-02	LIQUID	4.83E+3	4.64E+3	4.013 RPD	
4 DUP	S98T003010	0	@TICTOC1 TOC-02	LIQUID	2.38E+3	2.29E+3	3.854 RPD	
5 SAMPLE	S98T003011	0	@TICTOC1 TIC-02	LIQUID	N/A	4.57E+03	5.000	ug/mL
5 SAMPLE	S98T003011	0	@TICTOC1 TOC-02	LIQUID	N/A	2.33E+03	40.000	ug/mL
6 DUP	S98T003011	0	@TICTOC1 TIC-02	LIQUID	4.57E+3	4.45E+3	2.661 RPD	
6 DUP	S98T003011	0	@TICTOC1 TOC-02	LIQUID	2.33E+3	1.12E+3	70.145 RPD	
7 TRIPL	S98T003011	0	@TICTOC1 TIC-02	LIQUID	4.57E+3	4.49E+3	1.766 RPD	
7 TRIPL	S98T003011	0	@TICTOC1 TOC-02	LIQUID	2.33E+3	2.31E+3	0.862 RPD	

Final page for worklist# 26846

Analyst Signature _____ Date _____

Analyst Signature *CP Orsien* Date *11/25/98*

Reviewer Signature *[Signature]* Date *12/1/98*

10/27/98 07:18
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26846

Analyst: RL Instrument: CARB2 Book# TIC 25N12D
 Method: LA-342-100 Rev/Mod F-1 TOC 24 N12F

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	LIQUID		
2 STD			@TICTOC1	LIQUID		
3 SAMPLE	S98T003010 0		@TICTOC1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T003010 0		@TICTOC1	LIQUID		
5 SAMPLE	S98T003011 0		@TICTOC1	LIQUID	98000510	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
6 DUP	S98T003011 0		@TICTOC1	LIQUID		

Final page for worklist # 26846

TRIP

RJ McCown 11/18/98
 Analyst Signature Date

Ch. Quinn 11/25/98
 Analyst Signature Date

Data Entry Comments:

598T003011 run in triplicate

5450

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3691

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	5.8	28.9
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code			
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:45 PM		5	40
Sample Point	µg of Carbon	1.30E+00	1.70E+00
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>A/A</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3692

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	611.4	592.3
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000

Matrix
LIQUID

Batch Number
98005450

Rerun
0

Sample Prep
N/A

Sample #
STD

Instrument Code
CARB1

Prepared By
CJO

Chemist
MJL

Analyst
PJM

Date Complete
11/25/98

Analysis Date
11/11/98

Analysis Time
04:45 PM

Sample Point
BY-105

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
04:45 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.04E+02	2.83E+03
BY-105	Percent Standard Recovery	100.4	94.2

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3693

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	489.6	265.4
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
NA			
Sample #			
S98T003010			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:45 PM			
Sample Point	µg of Carbon/mL	4.83E+03	2.38E+03
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	471.1	256
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code	Known µg of C from Original Sample	4.83E+3	2.38E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003010			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:45 PM			
Sample Point	µg of Carbon/mL	4.64E+03	2.29E+03
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3695

WORKBOOK PAGE: SAM5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Size in mL (SS)	TIC	TOC
SAMPLE	Dilution Factor (DF)	0.1000	0.1000
Work List	μg of Carbon in Sample (C1)	463.6	260.4
26846	μg of Carbon from Baseline (C2)	7.1	27.2
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	μg of Carbon/mL = (C1-C2) * DF / SS		
0	μg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	μg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003011			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	TIC	TOC
04:45 PM		5	40
Sample Point	μg of Carbon/mL	4.57E+03	2.33E+03
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	451.6	139.4
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code	Known µg of C from Original Sample	4.57E+3	2.33E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003011			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:45 PM		5	40
Sample Point	µg of Carbon/mL	4.45E+03	1.12E+03
BY-105			

Data Entered By:	CJO	Date:	11/25/98
Signature of Chemist:	NA	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3697

WORKBOOK PAGE: TRIPL7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	456.1	258.7
26846	µg of Carbon from Baseline (C2)	7.1	27.2
Test Code	Known µg of C from Original Sample	4.57E+3	2.33E+3
@TICTOC1	Duplicate Result	4.45E+3	1.12E+3
Matrix			
LIQUID			
Batch Number			
98005450			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003011			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/25/98			
Analysis Date			
11/11/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:45 PM			
Sample Point	µg of Carbon/mL	4.49E+03	2.31E+03
BY-105			

Data Entered By:

CJO

Date:

11/25/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3698

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

Sample: STD-1

Date: 11/06/98

Time: 22:44:02

Sample Size = 1000 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .65 ug/minute C

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	0.40	0.00
2	1.01	5.40	92.59
3	1.51	95.10	94.32
4	2.00	270.00	64.78
5	2.50	432.60	37.59
6	3.00	529.00	18.22
7	3.50	575.70	8.11
8	4.00	595.50	3.32
9	4.50	602.90	1.23
10	5.00	605.40	0.41
11	5.50	606.80	0.23
12	6.00	607.60	0.13
13	6.50	608.00	0.07
14	7.00	608.50	0.08
15	7.50	608.90	0.07
16	8.00	609.30	0.07
17	8.50	609.80	0.08
18	9.00	610.00	0.03
19	9.50	610.40	0.07
20	10.00	610.70	0.05
21	10.50	611.00	0.05
22	11.00	611.40	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(611.4 - 7.148651) (1)/(1000) =

+6.043E-01 g/L Carbon

(611.4 - 7.148651) (1)/(1000) (12) =

+5.035E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3699 TO 3714

Sample Run By:

PJ MCCOWN

00000

1.0 mL 25N12D

3699

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: STD-1

Date: 11/06/98

Time: 22:58:20

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute.C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	1.20	75.00
3	1.50	8.30	85.54
4	2.00	61.80	86.57
5	2.50	212.20	70.88
6	3.00	417.40	49.16
7	3.50	530.70	21.35
8	4.00	568.10	6.58
9	4.50	580.40	2.12
10	5.00	584.50	0.70
11	5.50	586.20	0.29
12	6.00	587.50	0.22
13	6.50	588.10	0.10
14	7.00	588.90	0.14
15	7.50	589.20	0.05
16	8.00	589.90	0.12
17	8.50	590.30	0.07
18	9.00	590.80	0.08
19	9.50	591.20	0.07
20	10.00	591.80	0.10
21	10.50	592.00	0.03
22	11.00	592.30	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978

= +2.5E+00

ug/min Carbon

SAMPLE RESULTS:

(592.3 - 27.16457) (1)/(200) =

+2.826E+00 g/L Carbon

(592.3 - 27.16457) (1)/(200) (12) =

+2.355E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

200 mL 26 N12 F

3700

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-1

Date: 11/06/98

Time: 21:20:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-1

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	0.40	0.00
2	1.01	0.90	55.56
3	1.51	1.30	30.77
4	2.00	1.70	23.53
5	2.50	1.90	10.53
6	3.00	2.20	13.64
7	3.50	2.60	15.38
8	4.00	2.90	10.34
9	4.50	3.30	12.12
10	5.00	3.50	5.71
11	5.50	3.80	7.89
12	6.00	4.20	9.52
13	6.50	4.50	6.67
14	7.00	4.80	6.25
15	7.50	5.10	5.88
16	8.00	5.40	5.56
17	8.50	5.60	3.57
18	9.00	5.90	5.08
19	9.50	6.20	4.84
20	10.00	6.60	6.06
21	10.50	7.00	5.71
22	11.00	7.10	1.41

BLANK VALUE = 7.1 micrograms carbon

BLANK FACTOR = 7.1 / 10.99878 =

+6.5E-01

ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

3701

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV. 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-1

Date: 11/06/98

Time: 22:27:02

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-1

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.30	0.00
2	1.00	1.20	75.00
3	1.50	4.10	70.73
4	2.00	9.10	54.95
5	2.50	14.10	35.46
6	3.00	17.50	19.43
7	3.50	19.60	10.71
8	4.00	20.70	5.31
9	4.50	21.50	3.72
10	5.00	22.20	3.15
11	5.50	22.80	2.63
12	6.00	23.30	2.15
13	6.50	23.90	2.51
14	7.00	24.30	1.65
15	7.50	24.80	2.02
16	8.00	25.10	1.20
17	8.50	25.50	1.57
18	9.00	25.90	1.54
19	9.50	26.20	1.15
20	10.00	26.60	1.50
21	10.50	26.90	1.12
22	11.00	27.20	1.10

BLANK VALUE = 27.2 micrograms carbon

BLANK FACTOR = 27.2 / 10.9978 =

+2.47E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3702

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-1

Date: 11/06/98

Time: 23:13:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	0.80	37.50
3	1.50	1.10	27.27
4	2.00	1.50	26.67
5	2.50	1.60	6.25
6	3.00	2.00	20.00
7	3.50	2.10	4.76
8	4.00	2.40	12.50
9	4.50	2.70	11.11
10	5.00	2.90	6.90
11	5.50	3.10	6.45
12	6.00	3.50	11.43
13	6.50	3.70	5.41
14	7.00	3.90	5.13
15	7.50	4.10	4.88
16	8.00	4.40	6.82
17	8.50	4.80	8.33
18	9.00	4.90	2.04
19	9.50	5.30	7.55
20	10.00	5.40	1.85
21	10.50	5.70	5.26
22	11.00	5.80	1.72

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978 =

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(5.8 - 7.148571) (1)/(1) =

< 5.00 E-3 g/L Carbon

(5.8 - 7.148571) (1)/(1) (12) =

< 4.17 E-4 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

3703

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: BLNK-1

Date: 11/06/98

Time: 23:27:09

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.50	0.10	0.00
2	1.04	1.00	90.00
3	1.54	3.30	69.70
4	2.04	8.10	59.26
5	2.54	13.70	40.88
6	3.04	17.70	22.60
7	3.54	20.30	12.81
8	4.04	21.60	6.02
9	4.54	22.40	3.57
10	5.04	23.00	2.61
11	5.54	23.70	2.95
12	6.04	24.20	2.07
13	6.54	24.80	2.42
14	7.04	25.40	2.36
15	7.53	26.10	2.68
16	8.03	26.50	1.51
17	8.53	27.00	1.85
18	9.03	27.50	1.82
19	9.53	27.80	1.08
20	10.03	28.20	1.42
21	10.53	28.50	1.05
22	11.03	28.90	1.38

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978 = +2.5E+00 ug/min Carbon

SAMPLE RESULTS:

(28.9 - 27.24839) (1)/(1) = +1.65E+00 g/L Carbon
 (28.9 - 27.24839) (1)/(1) (12) = +1.38E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3704

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003010

Date: 11/06/98

Time: 23:45:15

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.00	0.70	57.14
3	1.50	1.60	95.21
4	2.00	92.60	84.23
5	2.50	236.40	60.83
6	3.00	360.40	34.41
7	3.50	431.80	16.54
8	4.00	463.20	6.78
9	4.50	475.30	2.55
10	5.00	480.40	1.06
11	5.50	482.70	0.48
12	6.00	483.90	0.25
13	6.50	484.90	0.21
14	7.00	485.60	0.14
15	7.50	486.20	0.12
16	8.00	486.80	0.12
17	8.50	487.30	0.10
18	9.00	487.80	0.10
19	9.50	488.10	0.06
20	10.00	488.70	0.12
21	10.50	489.10	0.08
22	11.00	489.60	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978 =

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(489.6 - 7.148571) (1)/(1) =

+4.825E+02 g/L Carbon

(489.6 - 7.148571) (1)/(1)(12) =

+4.020E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.100ML

3705

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003010

Date: 11/06/98

Time: 23:58:55

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.01	1.10	45.45
3	1.51	5.50	80.00
4	2.01	26.90	79.55
5	2.50	97.50	72.41
6	3.00	178.30	45.32
7	3.50	221.80	19.61
8	4.00	241.60	8.20
9	4.50	250.20	3.44
10	5.00	254.60	1.73
11	5.50	257.10	0.97
12	6.00	258.70	0.62
13	6.50	259.90	0.46
14	7.00	260.80	0.35
15	7.50	261.80	0.38
16	8.00	262.30	0.19
17	8.50	263.00	0.27
18	9.00	263.80	0.30
19	9.50	264.10	0.11
20	10.00	264.60	0.19
21	10.50	265.10	0.19
22	11.00	265.40	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978 = +2.5E+00 ug/min Carbon

SAMPLE RESULTS:

(265.4 - 27.16729) (1)/(1) = +2.382E+02 g/L Carbon
 (265.4 - 27.16729) (1)/(1)(12) = +1.985E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

600 mL

3706

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV:2.0

Sample: 3010 DUP

Date: 11/07/98

Time: 00:39:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.70	42.86
3	1.50	12.10	94.21
4	2.00	92.80	86.96
5	2.50	236.50	60.76
6	3.00	356.80	33.72
7	3.50	421.50	15.35
8	4.00	448.90	6.10
9	4.50	459.10	2.22
10	5.00	462.70	0.78
11	5.50	464.50	0.39
12	6.00	465.60	0.24
13	6.50	466.40	0.17
14	7.00	467.10	0.15
15	7.50	467.70	0.13
16	8.00	468.30	0.13
17	8.50	468.70	0.09
18	9.00	469.30	0.13
19	9.50	469.70	0.09
20	10.00	470.20	0.11
21	10.50	470.70	0.11
22	11.00	471.10	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(471.1 - 7.148698) (1)/(1) =

+4.640E+02 g/L Carbon

(471.1 - 7.148698) (1)/(1) (12) =

+3.866E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

100mL

3707

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3010 DUP

Date: 11/07/98

Time: 00:54:43

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst :

PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.01	1.00	40.00
3	1.50	5.60	82.14
4	2.00	26.00	78.46
5	2.50	98.30	73.55
6	3.00	177.30	44.56
7	3.50	217.80	18.60
8	4.00	234.10	6.96
9	4.50	242.10	3.30
10	5.00	245.60	1.43
11	5.50	248.00	0.97
12	6.00	249.70	0.68
13	6.50	250.90	0.48
14	7.00	251.80	0.36
15	7.50	252.40	0.24
16	8.00	253.10	0.28
17	8.50	253.70	0.24
18	9.00	254.20	0.20
19	9.50	254.70	0.20
20	10.00	255.10	0.16
21	10.50	255.60	0.20
22	11.00	256.00	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978 = +2.5E+00 ug/min Carbon

SAMPLE RESULTS:

(256 - 27.16506) (1) / (1) =

+2.288E+02 g/L Carbon

(256 - 27.16506) (1) / (1) (12) =

+1.907E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ McCown

PJ MCCOWN

00000

, 00 uL

3708

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV.2.0

Sample: S98T003011

Date: 11/07/98

Time: 01:41:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.01	1.20	66.67
3	1.50	24.80	95.16
4	2.00	120.30	79.38
5	2.50	258.00	53.37
6	3.00	365.50	29.41
7	3.50	420.60	13.10
8	4.00	443.20	5.10
9	4.50	452.20	1.99
10	5.00	455.60	0.75
11	5.50	457.20	0.35
12	6.00	458.20	0.22
13	6.50	459.00	0.17
14	7.00	459.70	0.15
15	7.50	460.30	0.13
16	8.00	460.80	0.11
17	8.50	461.30	0.11
18	9.00	461.80	0.11
19	9.50	462.40	0.13
20	10.00	462.80	0.09
21	10.50	463.10	0.06
22	11.00	463.60	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978 =

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(463.6 - 7.148592) (1)/(1) =

+4.565E+02 g/L Carbon

(463.6 - 7.148592) (1)/(1) (12) =

+3.804E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

3709

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2 0

Sample: S98T003011

Date: 11/07/98

Time: 01:55:10

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

```

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==
 1          0.51          0.20          0.00
 2          1.01          0.90          77.78
 3          1.51          5.50          83.64
 4          2.00          26.00          78.85
 5          2.50          94.20          72.40
 6          3.00          173.60          45.74
 7          3.50          216.80          19.93
 8          4.00          236.80          8.45
 9          4.50          245.30          3.47
10          5.00          249.30          1.60
11          5.50          252.20          1.15
12          6.00          253.50          0.51
13          6.50          254.90          0.55
14          7.00          256.20          0.51
15          7.50          256.80          0.23
16          8.00          257.50          0.27
17          8.50          258.10          0.23
18          9.00          258.60          0.19
19          9.50          259.10          0.19
20          10.00         259.70          0.23
21          10.50         260.00          0.12
22          11.00         260.40          0.15

```

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978 = +2.5E+00 ug/min Carbon

SAMPLE RESULTS:

```

( 260.4 - 27.16508 ) (1) / (1) = +2.332E+02 g/L Carbon
( 260.4 - 27.16508 ) (1) / (1) (12) = +1.944E+01 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

```

Sample Run By:

PJ MCCOWN

00000

-100mL

3710

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3011 DUP

Date: 11/07/98

Time: 02:08:43

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.01	1.00	50.00
3	1.50	5.50	81.82
4	2.00	36.50	84.93
5	2.50	107.10	65.92
6	3.00	232.70	53.98
7	3.50	347.00	32.94
8	4.00	404.20	14.15
9	4.50	429.40	5.87
10	5.00	439.80	2.36
11	5.50	443.80	0.90
12	6.00	445.60	0.40
13	6.50	446.50	0.20
14	7.00	447.30	0.18
15	7.50	448.00	0.16
16	8.00	448.60	0.13
17	8.50	449.20	0.13
18	9.00	449.70	0.11
19	9.50	450.30	0.13
20	10.00	450.80	0.11
21	10.50	451.30	0.11
22	11.00	451.60	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.148571 / 10.9978

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(451.6 - 7.148592) (1)/(1) =

+4.445E+02 g/L Carbon

(451.6 - 7.148592) (1)/(1).(12) =

+3.704E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.100mL

3711

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 3011 DUP

Date: 11/07/98

Time: 02:22:53

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.20	0.00
2	1.01	0.80	75.00
3	1.51	2.90	72.41
4	2.00	7.90	63.29
5	2.50	14.40	45.14
6	3.00	25.30	43.08
7	3.50	42.80	40.89
8	4.00	63.90	33.02
9	4.50	85.70	25.44
10	5.00	103.20	16.96
11	5.50	116.40	11.34
12	6.00	125.00	6.88
13	6.50	130.10	3.92
14	7.00	133.20	2.33
15	7.50	135.00	1.33
16	8.00	136.20	0.88
17	8.50	137.00	0.58
18	9.00	137.60	0.44
19	9.50	138.10	0.36
20	10.00	138.60	0.36
21	10.50	139.00	0.29
22	11.00	139.40	0.29

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978 = +2.5E+00 ug/min Carbon

SAMPLE RESULTS:

(139.4 - 27.1671) (1)/(1) = +1.122E+02 g/L Carbon
 (139.4 - 27.1671) (1)/(1) (12) = +9.353E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ

PJ MCCOWN

00000

.100mL

3712

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3011 TRIP

Date: 11/07/98

Time: 02:38:54

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.50	0.00
2	1.05	1.30	61.54
3	1.55	29.30	95.56
4	2.05	146.50	80.00
5	2.55	295.50	50.42
6	3.04	388.20	23.88
7	3.54	426.90	9.07
8	4.04	441.20	3.24
9	4.54	446.30	1.14
10	5.04	448.20	0.42
11	5.54	449.50	0.29
12	6.04	450.40	0.20
13	6.54	451.20	0.18
14	7.04	452.00	0.18
15	7.54	452.50	0.11
16	8.04	453.10	0.13
17	8.54	453.70	0.13
18	9.04	454.20	0.11
19	9.54	454.80	0.13
20	10.04	455.20	0.09
21	10.54	455.70	0.11
22	11.04	456.10	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 7.148571 micrograms carbon

BLANK FACTOR = 7.143571 / 10.9978 =

+6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(456.1 - 7.175351) (1)/(1) =

+4.489E+02 g/L Carbon

(456.1 - 7.175351) (1)/(1) (12) =

+3.741E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.100 mL

3713

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3011 TRIP

Date: 11/07/98

Time: 02:52:32

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.47 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.20	50.00
3	1.50	8.20	76.92
4	2.00	24.60	78.86
5	2.50	89.90	72.64
6	3.00	169.60	46.99
7	3.50	213.90	20.71
8	4.00	234.10	8.63
9	4.50	242.70	3.54
10	5.00	247.20	1.82
11	5.50	249.60	0.96
12	6.00	251.70	0.83
13	6.50	252.70	0.40
14	7.00	253.80	0.43
15	7.50	254.70	0.35
16	8.00	255.40	0.27
17	8.50	256.10	0.27
18	9.00	256.70	0.23
19	9.50	257.20	0.19
20	10.00	257.70	0.19
21	10.50	258.40	0.27
22	11.00	258.70	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 27.16457 micrograms carbon

BLANK FACTOR = 27.16457 / 10.9978

+2.5E+00

ug/min Carbon

SAMPLE RESULTS:

(258.7 - 27.16465) (1)/(1) =

+2.315E+02 g/L Carbon

(258.7 - 27.16465) (1)/(1)(12) =

+1.929E+01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ McCown

PJ MCCOWN

00000

100 mL

3714

LABCORE Completed Worklist Report for Worklist# 26855

Analyst: rdm Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	0	@TICTOC1 TIC-02	LIQUID	1	6.50E+0	6.500	ug/mL
1 BLNK	0	0	@TICTOC1 TOC-02	LIQUID	1	7.90E+0	7.900	ug/mL
2 STD	0	0	@TICTOC1 TIC-02	LIQUID	6.02E+02	5.71E+2	94.850	% Recovery
2 STD	0	0	@TICTOC1 TOC-02	LIQUID	3.00E+03	2.72E+3	90.667	% Recovery
3 SAMPLE	S98T003012	0	@TICTOC1 TIC-02	LIQUID	N/A	3.23E+03	5.000	ug/mL
3 SAMPLE	S98T003012	0	@TICTOC1 TOC-02	LIQUID	N/A	2.55E+03	40.000	ug/mL
4 DUP	S98T003012	0	@TICTOC1 TIC-02	LIQUID	3.23E+3	2.93E+3	9.740	RPD
4 DUP	S98T003012	0	@TICTOC1 TOC-02	LIQUID	2.55E+3	2.28E+3	11.180	RPD
5 SPK	S98T003012	0	@TICTOC1 TIC-02	LIQUID	1.00E+02	8.81E+01	88.100	% Recovery
5 SPK	S98T003012	0	@TICTOC1 TOC-02	LIQUID	1.00E+02	8.09E+01	80.900	% Recovery
6 SAMPLE	S98T003013	0	@TICTOC1 TIC-02	LIQUID	N/A	3.10E+03	5.000	ug/mL
6 SAMPLE	S98T003013	0	@TICTOC1 TOC-02	LIQUID	N/A	1.95E+03	40.000	ug/mL
7 DUP	S98T003013	0	@TICTOC1 TIC-02	LIQUID	3.10E+3	3.38E+3	8.642	RPD
7 DUP	S98T003013	0	@TICTOC1 TOC-02	LIQUID	1.95E+3	2.10E+3	7.407	RPD

Final page for worklist# 26855

Analyst Signature Date 11/10/98

Analyst Signature Date

Reviewer Signature Date 11/17/98

10/27/98 07:18
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26855

Analyst: RDM Instrument: CARB2 Book# LSMID 26N12FMethod: LA-342-100 Rev/Mod F-1

Worklist Comment: BY-105 FOR @TICTOC1 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	BLNK				@TICTOC1	LIQUID		
2	STD				@TICTOC1	LIQUID		
3	SAMPLE	S98T003012 0			@TICTOC1	LIQUID	98000510	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
4	DUP	S98T003012 0			@TICTOC1	LIQUID		
5	SPK	S98T003012 0			@TICTOC1	LIQUID		
6	SAMPLE	S98T003013 0			@TICTOC1	LIQUID	98000510	BY-105 (2)
		Analytes Requested: TIC-02 , TOC-02						
7	DUP	S98T003013 0			@TICTOC1	LIQUID		

Final page for worklist # 26855

RDM 11/8/98
 Analyst Signature Date

Ch. Quinn 11/16/98
 Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3716

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	8.9	30.8
26855	µg of Carbon from Baseline (C2)	2.4	38.7
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLK			
Instrument Code			
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:20 AM		5	40
Sample Point	µg of Carbon	6.50E+00	7.90E+00
BY105			

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK WB1 REV 1.0

342100ML

3717

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	573.4	583
26855	μg of Carbon from Baseline (C2)	2.4	38.7
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun			
0	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)		
Sample Prep	QC Found in $\mu\text{g}/\text{mL}$ = (C1 - C2) * DF / SS		
N/A	QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if C1 < C2		
Sample #	QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if C1 < C2		
STD			
Instrument Code	% Recovery = QC Found / QC Actual * 100		
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
11:20 AM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.71E+02	2.72E+03
BY105	Percent Standard Recovery	94.9	90.7

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>N/A</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3718

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	325.1	294.1
26855	µg of Carbon from Baseline (C2)	2.4	38.7
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003012			
Instrument Code			
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
11:20 AM			
Sample Point	µg of Carbon/mL	3.23E+03	2.55E+03
BY105			

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>N/A</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3719

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	295.7	266.5
26855	µg of Carbon from Baseline (C2)	2.4	38.7
Test Code	Known µg of C from Original Sample	3.23E+3	2.55E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003012			
Instrument Code			
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:20 AM		5	40
Sample Point	µg of Carbon/mL	2.93E+03	2.28E+03
BY105			

Data Entered By:

JDS

Date:

11/10/98

Signature of Chemist:

N/A

Date:

SAMPLE.WB1 REV 1.0

342100ML

3720

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Volume in mL (SS)	0.1000	0.1000
Work List	Final Coulometer Reading in µg (C1)	325.1	294.1
26855	Spiked Vial Data		
Test Code	Sample Volume in mL (SPK SS)	0.1000	0.1000
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	590.4	536.7
LIQUID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98005459	µg C in baseline (BL)	2.4	38.7
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T003012			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time			
11:20 AM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.31E+02	2.43E+03
BY105	Percent Spike Recovery	88.1	80.9

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>MA</i>	Date:	

SPIKE WB1 REV 1.3

342100ML

3721

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	312.6	233.5
26855	µg of Carbon from Baseline (C2)	2.4	38.7
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003013			
Instrument Code			
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:20 AM		5	40
Sample Point	µg of Carbon/mL	3.10E+03	1.95E+03
BY105			

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3722

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	340.4	248.5
26855	µg of Carbon from Baseline (C2)	2.4	38.7
Test Code	Known µg of C from Original Sample	3.10E+3	1.95E+3
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98005459			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003013			
Instrument Code			
CARB2			
Prepared By			
JDS			
Chemist			
MJL			
Analyst			
RDM			
Date Complete			
11/10/98			
Analysis Date			
11/08/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:20 AM		5	40
Sample Point	µg of Carbon/mL	3.38E+03	2.10E+03
BY105			

Data Entered By:	JDS	Date:	11/10/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3723

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
 TIC/OC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/05/98

Time: 23:26:39

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : RD MEYERS
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.00	0.00
2	1.01	0.10	100.00
3	1.51	0.20	50.00
4	2.00	0.40	50.00
5	2.50	0.60	33.33
6	3.00	0.70	14.29
7	3.50	0.90	22.22
8	4.00	1.00	10.00
9	4.50	1.10	9.09
10	5.00	1.20	8.33
11	5.50	1.30	7.69
12	6.00	1.40	7.14
13	6.50	1.50	6.67
14	7.00	1.60	6.25
15	7.50	1.70	5.88
16	8.00	1.80	5.56
17	8.50	1.80	0.00
18	9.00	2.00	10.00
19	9.50	2.00	0.00
20	10.00	2.10	4.76
21	10.50	2.20	4.55
22	11.00	2.40	8.33

BLANK VALUE = 2.4 micrograms carbon

BLANK FACTOR = 2.4 / 10.99792 = 0.2182 +2.2E-01 ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3724 TO 3732

Sample Run By:

RD MEYERS

00001

3724

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE

Date: 11/05/98

Time: 21:39:43

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE

Blank Value = N/A

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.10	0.00
2	1.01	1.00	90.00
3	1.51	35.40	84.38
4	2.00	14.00	54.29
5	2.50	21.50	34.88
6	3.00	26.10	17.62
7	3.50	28.90	9.69
8	4.00	30.90	6.47
9	4.50	32.30	4.33
10	5.00	33.50	3.58
11	5.50	34.50	2.90
12	6.00	35.00	1.43
13	6.50	35.70	1.96
14	7.00	36.30	1.65
15	7.50	36.80	1.36
16	8.00	37.00	0.54
17	8.50	37.30	0.80
18	9.00	37.50	0.53
19	9.50	37.80	0.79
20	10.00	38.10	0.79
21	10.50	38.50	1.04
22	11.00	38.70	0.52

BLANK VALUE = 38.7 micrograms Carbon

BLANK FACTOR = 38.7 / 10.99707 = +3.52E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 25N12-D

Date: 11/05/98

Time: 23:55:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.20	0.00
2	1.00	0.60	66.67
3	1.50	30.70	98.05
4	2.00	150.30	79.57
5	2.50	288.10	47.83
6	3.00	406.10	29.06
7	3.50	484.60	16.20
8	4.00	530.80	8.70
9	4.50	553.00	4.01
10	5.00	563.20	1.81
11	5.50	567.30	0.72
12	6.00	569.10	0.32
13	6.50	570.10	0.18
14	7.00	570.70	0.11
15	7.50	571.10	0.07
16	8.00	571.60	0.09
17	8.50	571.90	0.05
18	9.00	572.30	0.07
19	9.50	572.60	0.05
20	10.00	572.90	0.05
21	10.50	573.10	0.03
22	11.00	573.40	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(573.4 - 2.419355) (1)/(1) = +5.710E+02 g/L Carbon
 (573.4 - 2.419355) (1)/(1)(12) = +4.758E+01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 26N12-F

Date: 11/06/98

Time: 00:09:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute, C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.30	0.00
2	1.01	6.60	95.45
3	%-1438.49	114.50	94.24
4	%-1438.00	304.30	62.37
5	%-1437.50	474.40	35.86
6	%-1437.00	540.80	12.28
7	%-1436.50	563.20	3.98
8	%-1436.00	571.00	1.37
9	%-1435.50	574.20	0.56
10	%-1435.00	576.00	0.31
11	%-1434.50	577.20	0.21
12	%-1434.00	578.20	0.17
13	%-1433.50	578.70	0.09
14	%-1433.00	579.50	0.14
15	%-1432.50	580.00	0.09
16	%-1432.00	580.60	0.10
17	%-1431.50	580.90	0.05
18	%-1431.00	581.30	0.07
19	%-1430.50	581.80	0.09
20	%-1430.00	582.20	0.07
21	%-1429.50	582.60	0.07
22	%-1429.00	583.00	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(583 --5030.087) (1) / (1) = +5.613E+03 g/L Carbon
 (583 --5030.087) (1) / (1) (12) = +4.678E+02 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/06/98

Time: 00:21:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.40	0.00
2	1.00	0.80	50.00
3	1.50	2.60	69.23
4	2.01	4.00	35.00
5	2.51	4.80	16.67
6	3.01	5.40	11.11
7	3.51	5.60	3.57
8	4.00	6.00	6.67
9	4.50	6.30	4.76
10	5.00	6.40	1.56
11	5.50	6.70	4.48
12	6.00	6.90	2.90
13	6.50	7.10	2.82
14	7.00	7.30	2.74
15	7.50	7.50	2.67
16	8.00	7.80	3.85
17	8.50	7.90	1.27
18	9.00	8.20	3.66
19	9.50	8.40	2.38
20	10.00	8.50	1.18
21	10.50	8.70	2.30
22	11.00	8.90	2.25

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(8.9 - 2.42033) (1)/(1) = +6.5E+00 g/L Carbon
 (8.9 - 2.42033) (1)/(1)(12) = +5.4E-01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 11/06/98

Time: 00:34:08

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.20	0.00
2	1.01	0.80	75.00
3	1.50	7.30	89.04
4	2.00	14.40	49.31
5	2.50	19.30	25.39
6	3.00	22.00	12.27
7	3.50	23.80	7.56
8	4.00	25.00	4.80
9	4.50	25.90	3.47
10	5.00	26.50	2.26
11	5.50	27.30	2.93
12	6.00	27.60	1.09
13	6.50	28.00	1.43
14	7.00	28.40	1.41
15	7.50	28.70	1.05
16	8.00	29.20	1.71
17	8.50	29.40	0.68
18	9.00	29.70	1.01
19	9.50	30.00	1.00
20	10.00	30.30	0.99
21	10.50	30.50	0.66
22	11.00	30.80	0.97

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(30.8 - 38.70943) (1)/(1) = < 5.00 E-3 g/L Carbon
 (30.8 - 38.70943) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012

Date: 11/06/98

Time: 00:48:34

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.50	0.00
2	1.01	0.60	16.67
3	1.50	2.60	76.92
4	2.00	23.40	88.89
5	2.50	103.20	77.33
6	3.00	199.90	48.37
7	3.50	263.20	24.05
8	4.00	296.70	11.29
9	4.50	310.20	4.35
10	5.00	316.00	1.84
11	5.50	318.20	0.69
12	6.00	319.50	0.41
13	6.50	320.80	0.41
14	7.00	321.40	0.19
15	7.50	322.10	0.22
16	8.00	322.50	0.12
17	8.50	323.00	0.15
18	9.00	323.50	0.15
19	9.50	323.90	0.12
20	10.00	324.30	0.12
21	10.50	324.80	0.15
22	11.00	325.10	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(325.1 - 2.419339) (1)/(1) = +3.227E+02 g/L Carbon
 (325.1 - 2.419339) (1)/(1) (12) = +2.689E+01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012

Date: 11/06/98

Time: 01:00:54

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.10	63.64
3	1.50	20.00	94.50
4	2.00	85.10	76.50
5	2.50	185.80	54.20
6	3.00	243.90	23.82
7	3.50	267.00	8.65
8	4.00	275.70	3.16
9	4.50	280.10	1.57
10	5.00	283.10	1.06
11	5.50	285.30	0.77
12	6.00	286.90	0.56
13	6.50	288.00	0.38
14	7.00	288.90	0.31
15	7.50	289.80	0.31
16	8.00	290.80	0.34
17	8.50	291.40	0.21
18	9.00	292.20	0.27
19	9.50	292.60	0.14
20	10.00	293.20	0.20
21	10.50	293.80	0.20
22	11.00	294.10	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(294.1 - 38.7159) (1)/(1) = +2.554E+02 g/L Carbon
 (294.1 - 38.7159) (1)/(1) (12) = +2.128E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012DUP

Date: 11/06/98

Time: 01:14:12

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.50	2.80	67.86
4	2.00	26.50	89.43
5	2.50	101.00	73.76
6	3.00	188.80	46.50
7	3.50	244.60	22.81
8	4.00	272.80	10.34
9	4.50	283.70	3.84
10	5.00	288.10	1.53
11	5.50	289.70	0.55
12	6.00	290.80	0.38
13	6.50	291.60	0.27
14	7.00	292.20	0.21
15	7.50	292.70	0.17
16	8.00	293.20	0.17
17	8.50	293.60	0.14
18	9.00	294.10	0.17
19	9.50	294.40	0.10
20	10.00	294.90	0.17
21	10.50	295.20	0.10
22	11.00	295.70	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(295.7 - 2.419742) (1)/(1)	=	+2.933E+02	g/L Carbon
(295.7 - 2.419742) (1)/(1) (12)	=	+2.444E+01	Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012DUP

Date: 11/06/98

Time: 01:26:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.70	76.47
3	1.50	23.70	92.83
4	2.00	92.50	74.38
5	2.50	179.50	48.47
6	3.00	224.80	20.15
7	3.50	242.60	7.34
8	4.00	250.20	3.04
9	4.50	254.30	1.61
10	5.00	256.60	0.90
11	5.50	258.40	0.70
12	6.00	259.60	0.46
13	6.50	260.70	0.42
14	7.00	262.00	0.50
15	7.50	263.00	0.38
16	8.00	263.40	0.15
17	8.50	264.00	0.23
18	9.00	264.50	0.19
19	9.50	265.00	0.19
20	10.00	265.50	0.19
21	10.50	266.10	0.23
22	11.00	266.50	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(266.5 - 38.70942) (1)/(1) = +2.278E+02 g/L Carbon
 (266.5 - 38.70942) (1)/(1)(12) = +1.898E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012SPK

Date: 11/06/98

Time: 01:40:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	0.90	44.44
3	1.50	6.90	86.96
4	2.00	93.20	92.60
5	2.50	244.10	61.82
6	3.00	387.10	36.94
7	3.50	484.50	20.10
8	4.00	540.10	10.29
9	4.50	566.30	4.63
10	5.00	577.20	1.89
11	5.50	581.60	0.76
12	6.00	583.70	0.36
13	6.50	585.00	0.22
14	7.00	586.00	0.17
15	7.50	586.70	0.12
16	8.00	587.20	0.09
17	8.50	587.90	0.12
18	9.00	588.40	0.08
19	9.50	589.00	0.10
20	10.00	589.50	0.08
21	10.50	590.00	0.08
22	11.00	590.40	0.07

USER INPUT BLANK VALUE:

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(590.4 - 2.41956) (1)/(1) = +5.880E+02 g/L Carbon
 (590.4 - 2.41956) (1)/(1) (12) = +4.900E+01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3012SPK

Date: 11/06/98

Time: 01:53:37

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	2.70	88.89
3	1.51	51.80	94.79
4	2.01	181.10	71.40
5	2.51	363.80	50.22
6	3.01	460.50	21.00
7	3.51	499.40	7.79
8	4.00	515.50	3.12
9	4.50	522.20	1.28
10	5.00	525.80	0.68
11	5.50	527.70	0.36
12	6.00	529.30	0.30
13	6.50	530.60	0.25
14	7.00	531.70	0.21
15	7.50	532.70	0.19
16	8.00	533.40	0.13
17	8.50	534.10	0.13
18	9.00	534.60	0.09
19	9.50	535.20	0.11
20	10.00	535.80	0.11
21	10.50	536.40	0.11
22	11.00	536.70	0.06

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(536.7 - 38.72236) (1)/(1) = +4.980E+02 g/L Carbon
 (536.7 - 38.72236) (1)/(1) (12) = +4.150E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3013

Date: 11/06/98

Time: 02:09:47

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.01	1.00	40.00
3	1.51	3.00	66.67
4	2.00	26.30	88.59
5	2.50	104.90	74.93
6	3.00	191.10	45.11
7	3.50	250.30	23.65
8	4.00	283.40	11.68
9	4.50	297.90	4.87
10	5.00	303.80	1.94
11	5.50	306.10	0.75
12	6.00	307.50	0.46
13	6.50	308.30	0.26
14	7.00	308.90	0.19
15	7.50	309.60	0.23
16	8.00	310.10	0.16
17	8.50	310.60	0.16
18	9.00	311.10	0.16
19	9.50	311.50	0.13
20	10.00	311.80	0.10
21	10.50	312.20	0.13
22	11.00	312.60	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(312.6 - 2.419926) (1)/(1) = +3.102E+02 g/L Carbon
 (312.6 - 2.419926) (1)/(1)(12) = +2.585E+01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2 0

Sample: 3013

Date: 11/06/98

Time: 01:23:19

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.30	0.00
2	1.01	0.70	57.14
3	1.50	10.30	93.20
4	2.00	49.80	79.32
5	2.50	124.90	60.13
6	3.00	181.30	31.11
7	3.50	206.70	12.29
8	4.00	216.90	4.70
9	4.50	221.40	2.03
10	5.00	224.50	1.38
11	5.50	226.20	0.75
12	6.00	227.50	0.57
13	6.50	228.40	0.39
14	7.00	229.20	0.35
15	7.50	229.80	0.26
16	8.00	230.60	0.35
17	8.50	231.30	0.30
18	9.00	231.90	0.26
19	9.50	232.40	0.22
20	10.00	232.80	0.17
21	10.50	233.20	0.17
22	11.00	233.50	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(233.5 - 38.71882) (1)/(1) = +1.948E+02 g/L Carbon
 (233.5 - 38.71882) (1)/(1) (12) = +1.623E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3013DUP

Date: 11/06/98

Time: 04:49:35

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .22 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.60	33.33
3	1.51	1.30	53.85
4	2.01	8.70	85.06
5	2.51	54.30	83.98
6	3.01	143.60	62.19
7	3.51	231.70	38.02
8	4.01	286.60	19.16
9	4.51	314.00	8.73
10	5.01	326.00	3.68
11	5.51	331.10	1.54
12	6.01	333.60	0.75
13	6.51	334.90	0.39
14	7.01	335.70	0.24
15	7.51	336.60	0.27
16	8.00	337.20	0.18
17	8.51	337.70	0.15
18	9.01	338.40	0.21
19	9.51	338.90	0.15
20	10.00	339.30	0.12
21	10.50	339.90	0.18
22	11.00	340.40	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 2.419355 micrograms carbon

BLANK FACTOR = 2.419355 / 10.99707 = +2.2E-01 ug/min Carbon

SAMPLE RESULTS:

(340.4 - 2.420953) (1)/(1) = +3.380E+02 g/L Carbon
 (340.4 - 2.420953) (1)/(1) (12) = +2.816E+01 Molar Carbon

Sample Run By:

RD MEYERS

00001

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: 3013DUP

Date: 11/06/98

Time: 05:02:51

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.52 ug/minute C

Analyst : RD MEYERS

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.01	1.30	53.85
3	1.50	20.90	93.78
4	2.00	89.70	76.70
5	2.50	175.00	48.74
6	3.00	213.40	17.99
7	3.50	227.90	6.36
8	4.00	234.20	2.69
9	4.50	239.50	1.39
10	5.00	239.50	0.84
11	5.50	240.90	0.58
12	6.00	242.60	0.70
13	6.50	243.50	0.37
14	7.00	244.30	0.33
15	7.50	244.90	0.24
16	8.00	245.50	0.24
17	8.50	246.20	0.28
18	9.00	246.80	0.24
19	9.50	247.20	0.16
20	10.00	247.60	0.16
21	10.50	248.20	0.24
22	11.00	248.50	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 38.70969 micrograms carbon

BLANK FACTOR = 38.70969 / 10.99707 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(248.5 - 38.71291) (1)/(1) = +2.098E+02 g/L Carbon
 (248.5 - 38.71291) (1)/(1)(12) = +1.748E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

RD MEYERS

00001

3739

LABCORE Completed Worklist Report for Worklist# 27218

Book#: _____

Method: LA-342-100   Rev/Mod _____

Worklist Comment: BY105 FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	0	ETICTOCI TIC-02	LIQUID	1	1.10E+0	1.100	ug/mL
1 BLNK	0	0	ETICTOCI TOC-02	LIQUID	1	3.30E+0	3.300	ug/mL
2 STD	0	0	ETICTOCI TIC-02	LIQUID	6.02E+02	6.06E+2	100.564	% Recovery
2 STD	0	0	ETICTOCI TOC-02	LIQUID	3.00E+03	2.85E+3	95.000	% Recovery
3 SAMPLE	S98T002966	0	ETICTOCI TIC-02	LIQUID	N/A	6.33E+03	5.000	ug/mL
3 SAMPLE	S98T002966	0	ETICTOCI TOC-02	LIQUID	N/A	2.28E+03	40.000	ug/mL
4 DUP	S98T002966	0	ETICTOCI TIC-02	LIQUID	6.33E+3	6.24E+3	1.432	RPD
4 DUP	S98T002966	0	ETICTOCI TOC-02	LIQUID	2.28E+3	2.17E+3	4.944	RPD

Final page for worklist# 27218

Analyst Signature _____ **Date** _____

Reviewer Signature _____ Date _____

11/16/98 15:03
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27218

Analyst: RLM Instrument: CARB2Book# TIC 25N12D
TOC 26N12FMethod: LA-342-100 Rev/Mod F-1

Worklist Comment: BY105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	LIQUID		
2 STD			@TICTOC1	LIQUID		
3 SAMPLE	S98T002966 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
			Analytes Requested: TIC-02 , TOC-02			
4 DUP	S98T002966 0		@TICTOC1	LIQUID		

Final page for worklist # 27218

RLM 11/31/98
Analyst Signature Date_____
Analyst Signature DateData Entry Comments:
_____
_____

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	7.8	26.5
27218	µg of Carbon from Baseline (C2)	6.7	29.8

Test Code
@TICTOC1
Matrix
LIQUID
Batch Number
98005830
Rerun
0
Sample Prep
N/A
Sample #
BLK
Instrument Code
CARB2
Prepared By
JRO
Chemist
MJL
Analyst
PJM
Date Complete
11/24/98
Analysis Date
11/21/98
Analysis Time
04:00 AM
Sample Point
BY-105

µg of Carbon = |C1-C2|

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 AM			
Sample Point	µg of Carbon	1.10E+00	3.30E+00
BY-105			

Data Entered By:	JRO	Date:	11/24/98
Signature of Chemist:		Date:	

BLANK.WB1 REV 1.0

342100ML

3742

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	612.3	600.4
27218	μg of Carbon from Baseline (C2)	6.7	29.8
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{ml}$)	602	3000
Matrix			
LIQUID			
Batch Number			
98005830			
Rerun			
0	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)		
Sample Prep	QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$		
N/A	QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$		
Sample #	QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$		
Instrument Code	% Recovery = QC Found / QC Actual * 100		
CARB2			
Prepared By			
JRO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/24/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
04:00 AM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.06E+02	2.85E+03
BY-105	Percent Standard Recovery	100.6	95.1

Data Entered By:	JRO	Date:	11/24/98
Signature of Chemist:	<i>NA</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3743

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	639.3	257.5
27218	µg of Carbon from Baseline (C2)	6.7	29.8
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005830			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002966			
Instrument Code			
CARB2			
Prepared By			
JRO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/24/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 AM			
Sample Point	µg of Carbon/mL	6.33E+03	2.28E+03
BY-105			

Data Entered By:

JRO

Date:

11/24/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	630.4	246.6
27218	µg of Carbon from Baseline (C2)	6.7	29.8
Test Code	Known µg of C from Original Sample	6.33E+3	2.28E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98005830			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002966			
Instrument Code			
CARB2			
Prepared By			
JRO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
11/24/98			
Analysis Date			
11/21/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:00 AM			
Sample Point	µg of Carbon/mL	6.24E+03	2.17E+03
BY-105			

Data Entered By:	JRO	Date:	11/24/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3745

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

Sample: STD-2

Date: 11/21/98

Time: 02:00:32

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .64 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.90	100.00
3	1.01	1.60	43.75
4	1.50	1.30	97.76
5	2.00	260.30	72.61
6	2.50	414.40	37.19
7	3.00	511.30	18.95
8	3.50	562.80	9.15
9	4.00	587.40	4.19
10	4.50	597.90	1.76
11	5.00	603.30	0.90
12	5.50	605.50	0.36
13	6.00	606.70	0.20
14	6.50	607.80	0.18
15	7.00	608.70	0.15
16	7.50	609.50	0.13
17	8.00	609.90	0.07
18	8.50	610.60	0.10
19	9.00	611.00	0.08
20	9.50	611.40	0.07
21	10.00	612.00	0.10
22	10.50	612.30	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 6.71936 micrograms carbon

BLANK FACTOR = 6.71936 / 10.499 = +6.4E-01 ug/min Carbon

SAMPLE RESULTS:

(612.3 - 6.71936) (1) / (1000) = +6.056E-01 g/L Carbon
 (612.3 - 6.71936) (1) / (1000) (12) = +5.047E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3744 TO 3755.

Sample Run By: PJ MCCOWN 11/21/98

PJ MCCOWN

00000

1.0 mL 25012 D

3746

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 11/21/98

Time: 02:14:13

Sample Size = 200 uL

Analyst : PJ MCCOWN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.84 ug/minute C

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	100.30	100.00
3	1.01	93.02	93.02
4	1.50	86.17	86.17
5	2.00	71.57	71.57
6	2.50	56.93	56.93
7	3.00	38.68	38.68
8	3.50	19.07	19.07
9	4.00	8.25	8.25
10	4.50	3.31	3.31
11	5.00	1.50	1.50
12	5.50	0.78	0.78
13	6.00	0.42	0.42
14	6.50	0.22	0.22
15	7.00	0.20	0.20
16	7.50	0.15	0.15
17	8.00	0.10	0.10
18	8.50	0.15	0.15
19	9.00	0.13	0.13
20	9.50	0.10	0.10
21	10.00	0.08	0.08
22	10.50	0.13	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 29.81716 micrograms carbon

BLANK FACTOR = 29.81716 / 10.499 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(600.4 - 29.81764) (1)/(200) = -2.853E+00 g/L Carbon

(600.4 - 29.81764) (1)/(200)(12) = +2.377E-01 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

.200ml 26012 F

3747

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/21/98

Time: 01:20:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-2

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading == Analysis Time == Coulometer == % Difference ==

1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	0.70	42.86
4	1:50	TIC: INORGANIC CARBON ANALYSIS	41.67
5	2:00	TICTOC REV 2.0	14.29
6	2:50	BLANK ANALYSIS	22.22
7	3:00	2.20	18.18
8	3:50	Date: 11/21/98	8.33
9	4:00	2.70	11.11
10	4:50	3.00 A	10.00
11	5:00	3.40 M	11.76
12	5:50	3.70 Max	8.11
13	6:00	3.90 %	5.13
14	6:50	4.20	7.14
15	7:00	Line from Coulometer	6.67
16	7:50	4.80	6.25
17	8:00	5.20	7.69
18	8:50	5.50	5.45
19	9:00	INORGANIC CARBON ANALYSIS	5.17
20	9:50	TICTOC REV 2.0	4.92
21	10:00	BLANK ANALYSIS	4.69
22	10:50	6.70	4.48

BLANK VALUE = 6.7 micrograms carbon

BLANK FACTOR = 6.7 / 10.5 = 0.638 +6.4E-01

ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

3748

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2-0

<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 11/21/95

Time: 01:39:51

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-2

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis	Time	Coulometer	% Difference
1	0.88	0:00	0.00	0.00
2	0.51	0:40	100.00	100.00
3	1.01	1:10	63.64	63.64
4	1:50	5:30	79.25	79.25
5	2:00	1:00	51.82	51.82
6	2:50	16:50	33.33	33.33
7	3:00	20:00	17.50	17.50
8	3:50	21:30	8.68	8.68
9	4:00	23:30	6.01	6.01
10	4:50	24:30	4.12	4.12
11	5:00	25:10	3.19	3.19
12	5:50	25:60	1.95	1.95
13	6:00	26:20	2.29	2.29
14	6:50	28:60	1.50	1.50
15	7:00	27:30	2.56	2.56
16	7:50	27:50	0.73	0.73
17	8:00	27:90	1.43	1.43
18	8:50	28:50	2.11	2.11
19	9:00	28:50	0.35	0.35
20	9:50	29:00	1.38	1.38
21	10:00	29:30	1.02	1.02
22	10:50	29:80	1.68	1.68

BLANK VALUE = 29.8 micrograms Carbon

BLANK FACTOR = 29.8 / 10.499 = 2.84E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 11/21/98

Time: 02:30:55

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .64 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	0.80	37.50
4	1.51	1.70	52.94
5	2.00	1.90	10.53
6	2.50	2.20	13.64
7	3.00	2.50	12.00
8	3.50	3.10	19.35
9	4.00	3.20	3.13
10	4.50	3.80	15.79
11	5.00	4.00	5.00
12	5.50	4.30	6.98
13	6.00	4.60	6.52
14	6.50	5.10	9.80
15	7.00	5.70	10.53
16	7.50	6.20	8.06
17	8.00	6.50	4.62
18	8.50	6.60	1.52
19	9.00	6.90	4.35
20	9.50	7.20	4.17
21	10.00	7.60	5.26
22	10.50	7.80	2.56

USER INPUT BLANK VALUE

BLANK VALUE = 6.71936 micrograms carbon

BLANK FACTOR = 6.71936 / 10.499 = +6.4E-01 ug/min Carbon

SAMPLE RESULTS:

(7.8 - 6.718291) (1) / (1) =

+1.1E+00

g/L Carbon

(7.8 - 6.718291) (1) / (1) (12) =

-9.0E-02

Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 11/21/98

Time: 02:44:01

Sample Size = 1 uL

Dil Factor = 1.

Blank ID # =

Blank Value = 2.84 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.70	57.14
4	1.50	1.00	30.00
5	2.00	1.70	41.18
6	2.50	3.50	51.43
7	3.00	7.00	50.00
8	3.50	10.60	33.96
9	4.00	14.90	28.86
10	4.50	17.20	13.37
11	5.00	19.20	10.42
12	5.50	20.40	5.88
13	6.00	21.70	5.99
14	6.50	22.60	3.98
15	7.00	23.10	2.16
16	7.50	23.80	2.94
17	8.00	24.20	1.65
18	8.50	24.70	2.02
19	9.00	25.50	3.14
20	9.50	25.80	1.16
21	10.00	26.30	1.90
22	10.50	26.50	0.75

USER INPUT BLANK VALUE

BLANK VALUE = 29.81716 micrograms carbon

BLANK FACTOR = 29.81716 / 10.499 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(26.5 - 29.81718) (1) / (1) = A 5.00 E-3 g/L Carbon
 (26.5 - 29.81718) (1) / (1) (12) = A 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3751

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002966

Date: 11/21/98

Time: 02:55:43

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .64 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	0.90	66.67
4	1.50	1.10	18.18
5	2.00	1.60	31.25
6	2.50	16.80	90.48
7	3.00	93.90	82.11
8	3.50	216.10	56.55
9	4.00	342.10	36.83
10	4.50	447.30	23.52
11	5.00	520.80	14.11
12	5.50	567.90	8.29
13	6.00	596.80	4.84
14	6.50	612.60	2.58
15	7.00	623.40	1.73
16	7.50	628.10	0.75
17	8.00	631.60	0.55
18	8.50	634.30	0.43
19	9.00	635.80	0.24
20	9.50	637.30	0.24
21	10.00	638.30	0.16
22	10.50	639.30	0.16

USER INPUT BLANK VALUE

BLANK VALUE = 6.71936 micrograms carbon

BLANK FACTOR = 6.71936 / 10.499 = +6.4E-01 ug/min Carbon

SAMPLE RESULTS:

(639.3 - 6.718818) (1)/(1) = +6.326E+02 g/L Carbon
 (639.3 - 6.718818) (1)/(1)(12) = +5.272E+01 Molar Carbon

Sample Run By: RML
 PJ MCCOWN

00000

3752

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002966

Date: 11/21/98

Time: 00:08:01

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.84 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.40	50.00
4	1.50	2.00	30.00
5	2.00	3.00	33.33
6	2.50	10.60	71.70
7	3.00	38.60	72.54
8	3.50	90.70	57.44
9	4.00	148.00	38.72
10	4.50	189.50	21.90
11	5.00	213.20	11.12
12	5.50	228.50	6.70
13	6.00	236.90	3.55
14	6.50	243.10	2.55
15	7.03	247.40	1.74
16	7.53	249.90	1.00
17	8.03	251.60	0.68
18	8.53	253.30	0.67
19	9.03	254.60	0.51
20	9.53	255.70	0.43
21	10.03	256.60	0.35
22	10.53	257.50	0.35

USER INPUT BLANK VALUE

BLANK VALUE = 29.81716 micrograms carbon

BLANK FACTOR = 29.81716 / 10.499 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(257.5 - 29.91087) (1)/(1) = +2.276E+02 g/L Carbon
 (257.5 - 29.91087) (1)/(1) (12) = +1.897E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN

PJ MCCOWN

00000

.100mL

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2966 DUP

Date: 11/21/98

Time: 03:20:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .64 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	1.60	50.00
4	1.50	2.10	23.81
5	2.00	5.40	61.11
6	2.50	55.60	90.29
7	3.00	171.50	67.58
8	3.50	308.50	44.41
9	4.00	428.80	28.06
10	4.50	508.60	15.69
11	5.00	560.90	9.32
12	5.50	590.90	5.08
13	6.00	606.50	2.57
14	6.50	614.80	1.35
15	7.00	619.70	0.79
16	7.50	622.80	0.50
17	8.00	625.30	0.40
18	8.50	626.40	0.18
19	9.00	627.70	0.21
20	9.50	629.00	0.21
21	10.00	629.70	0.11
22	10.50	630.40	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 6.71936 micrograms carbon

BLANK FACTOR = 6.71936 / 10.499 = +6.4E-01 ug/min Carbon

SAMPLE RESULTS:

(630.4 - 6.719355) (1)/(1) = +6.237E+02 g/L Carbon
 (630.4 - 6.719355) (1)/(1)(12) = +5.197E+01 Molar Carbon

Sample Run By: *R. McCown*

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2966 DUP

Date: 11/21/98

Time: 03:32:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.84 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	1.40	42.86
4	1.51	2.00	30.00
5	2.00	3.00	33.33
6	2.50	11.30	73.45
7	3.00	36.70	69.21
8	3.50	87.00	57.82
9	4.00	141.20	38.39
10	4.50	179.30	21.25
11	5.00	204.00	12.11
12	5.50	218.40	6.59
13	6.00	227.10	3.83
14	6.50	233.10	2.57
15	7.00	236.70	1.52
16	7.50	239.40	1.13
17	8.00	241.30	0.79
18	8.50	242.90	0.66
19	9.00	244.00	0.45
20	9.50	245.10	0.45
21	10.00	246.00	0.37
22	10.50	246.60	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 29.81716 micrograms carbon

BLANK FACTOR = 29.81716 / 19.499 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(246.6 - 29.81762) (1) / (1) = +2.168E+02 g/L Carbon
 (246.6 - 29.81762) (1) / (1) (12) = +1.807E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3755

LABCORE Completed Worklist Report for Worklist# 27371

Analyst: jds Instrument: TOC01 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105, @TICTOC1, ReRun at <0.1g sample size. cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	@TICTOC1 TIC-02	SOLID	6.02E+02	5.99E+2	99.502 % Recovery	
1 STD	0	0	@TICTOC1 TOC-02	SOLID	3.00E+03	2.75E+3	91.667 % Recovery	
2 BLNK	0	0	@TICTOC1 TIC-02	SOLID	1	1.10E+0	1.100 ug/g	
2 BLNK	0	0	@TICTOC1 TOC-02	SOLID	1	7.20E+0	7.200 ug/g	
3 SAMPLE	S98T002711	0	@TICTOC1 TIC-02	SOLID	N/A	3.12E+04	5.000 ug/g	
3 SAMPLE	S98T002711	0	@TICTOC1 TOC-02	SOLID	N/A	6.23E+03	40.000 ug/g	
4 DUP	S98T002711	0	@TICTOC1 TIC-02	SOLID	3.12E+4	2.44E+4	24.460 RPD	
4 DUP	S98T002711	0	@TICTOC1 TOC-02	SOLID	6.23E+3	5.75E+3	8.013 RPD	
5 SPK	S98T002711	0	@TICTOC1 TIC-02	SOLID	1.00E+02	0.00E+00	0.000 % Recovery	
5 SPK	S98T002711	0	@TICTOC1 TOC-02	SOLID	1.00E+02	2.02E+02	202.000 % Recovery	
6 SAMPLE	S98T002840	0	@TICTOC1 TIC-02	SOLID	N/A	2.47E+03	5.000 ug/g	
6 SAMPLE	S98T002840	0	@TICTOC1 TOC-02	SOLID	N/A	2.49E+03	40.000 ug/g	
7 DUP	S98T002840	0	@TICTOC1 TIC-02	SOLID	2.47E+3	2.98E+3	18.716 RPD	
7 DUP	S98T002840	0	@TICTOC1 TOC-02	SOLID	2.49E+3	2.14E+3	15.119 RPD	
8 SPK	S98T002840	0	@TICTOC1 TIC-02	SOLID	1.00E+02	1.88E+02	188.000 % Recovery	
8 SPK	S98T002840	0	@TICTOC1 TOC-02	SOLID	1.00E+02	9.98E+01	99.800 % Recovery	

Final page for worklist# 27371

Analyst Signature _____ Date _____

Manoj Jais 12-7-98
Analyst Signature _____ Date _____

MLA 12/8/98
Reviewer Signature _____ Date _____

12/01/98 15:34
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27371

Analyst: Jds Instrument: TOC01 Book# 25N12D
26N12F

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105, @TICTOC1, ReRun at <0.1g sample size. cjo

S Type	Sample#	R A Test	Matrix	Group#	Project
1 STD		@TICTOC1	SOLID		
2 BLNK		@TICTOC1	SOLID		
3 SAMPLE	S98T002711 0	@TICTOC1	SOLID	98000436	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02				
4 DUP	S98T002711 0	@TICTOC1	SOLID		
5 SPK	S98T002711 0	@TICTOC1	SOLID		
6 SAMPLE	S98T002840 0	@TICTOC1	SOLID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02				
7 DUP	S98T002840 0	@TICTOC1	SOLID		
8 SPK	S98T002840 0	@TICTOC1	SOLID		

Final page for worklist # 27371

Jds 12/5/98
 Analyst Signature Date

Marie Franz 12-7-98
 Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3757

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	610.5	589.1
27371	μg of Carbon from Baseline (C2)	11.8	39.5
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98005985			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
04:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.99E+02	2.75E+03
BY 105	Percent Standard Recovery	99.5	91.6

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	NA	Date:	
STANDARD.WB1 REV 1.0	342100ML		

3758

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	10.7	32.3
27371	µg of Carbon from Baseline (C2)	11.8	39.5
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005985			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in ug/g	5	40
04:00 PM			
Sample Point	µg of Carbon	1.10E+00	7.20E+00
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>MF</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3759

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0333	0.0333
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	1050.3	246.8
27371	µg of Carbon from Baseline (C2)	11.8	39.5
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98005985			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002711			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	3.12E+04	6.23E+03
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3760

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0369	0.0369
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	912.4	251.8
27371	µg of Carbon from Baseline (C2)	11.8	39.5
Test Code	Known µg of C from Original Sample	3.12E+4	6.23E+3
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98005985			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002711			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	2.44E+04	5.75E+03
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0333	0.0333
Work List	Final Coulometer Reading in µg (C1)	1050.3	246.8
27371	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0442	0.0442
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	1155.6	920.3
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98005985	µg C in baseline (BL)	11.8	39.5
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002711			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
TOC01	QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time		TIC	TOC
04:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	-4.69E+02	6.06E+03
BY 105	Percent Spike Recovery	-77.9	201.9

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3762

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0945	0.0945
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	245.4	274.8
27371	µg of Carbon from Baseline (C2)	11.8	39.5
Test Code			
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98005985			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002840			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	2.47E+03	2.49E+03
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3763

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.1339	0.1339
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	410.7	325.4
27371	µg of Carbon from Baseline (C2)	11.8	39.5
Test Code	Known µg of C from Original Sample	2.47E+3	2.49E+3
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98005985			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002840			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
04:00 PM		5	40
Sample Point	µg of Carbon/g	2.98E+03	2.14E+03
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3764

I:\342100\OUT\27371.WB1

12/07/98

10:04:04

WORKBOOK PAGE: SPIKE8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0945	0.0945
Work List	Final Coulometer Reading in μg (C1)	245.4	274.8
27371	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0830	0.0830
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	1.000	0.100
Matrix	Final Coulometer Reading in μg (C2)	1349	545.7
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in $\mu\text{g}/\text{mL}$ (SPK CONC)	602	3000
98005985	μg C in baseline (BL)	11.8	39.5
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = $((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100$		
S98T002840			
Instrument Code	QC Actual in $\mu\text{g}/\text{mL}$ = Spike Value ($\mu\text{g}/\text{mL}$)		
TOC01	QC Found in $\mu\text{g}/\text{mL}$ = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/05/98			
Analysis Time			
04:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	1.13E+03	3.00E+03
BY 105	Percent Spike Recovery	188.0	99.8

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	NA	Date:	

SPIKE.WB1 REV 1.3

342100ML

3765

X

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE

Date: 12/05/98

Time: 07:26:46

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : JD SPELLMAN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.90	100.00
3	1.01	1.50	40.00
4	1.50	2.20	31.82
5	2.00	2.90	24.14
6	2.50	3.60	19.44
7	3.00	4.30	16.28
8	3.50	4.80	10.42
9	4.00	5.40	11.11
10	4.50	6.00	10.00
11	5.00	6.60	9.09
12	5.50	6.90	4.35
13	6.00	7.40	6.76
14	6.50	7.70	3.90
15	7.00	8.20	6.10
16	7.50	8.80	6.82
17	8.00	9.10	3.30
18	8.50	9.70	6.19
19	9.00	10.00	3.00
20	9.50	10.50	4.76
21	10.00	11.10	5.41
22	10.50	11.80	5.93

BLANK VALUE = 11.8 micrograms carbon
 BLANK FACTOR = 11.8 / 10.49799 =

+1.12E+00 ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3766 TO 3782.

Sample Run By:

JD SPELLMAN

22222

3766

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE

Date: 12/05/98

Time: 07:42:03

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : JD SPELLMAN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.30	100.00
3	1.01	2.60	50.00
4	1.51	6.20	58.06
5	2.00	13.20	53.03
6	2.50	19.60	32.65
7	3.00	21.50	8.84
8	3.50	23.80	9.66
9	4.00	25.40	6.30
10	4.50	27.10	6.27
11	5.00	28.70	5.57
12	5.50	29.90	4.01
13	6.00	31.20	4.17
14	6.50	32.20	3.11
15	7.00	33.40	3.59
16	7.50	34.20	2.34
17	8.00	35.10	2.56
18	8.50	36.10	2.77
19	9.00	36.90	2.17
20	9.50	37.90	2.64
21	10.00	38.60	1.81
22	10.50	39.50	2.28

BLANK VALUE = 39.5 micrograms carbon

BLANK FACTOR = 39.5 / 10.49902 =

+3.76E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3767

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 12/05/98

Time: 07:56:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	2.10	61.90
4	1.51	6.20	66.13
5	2.00	135.60	95.43
6	2.50	337.60	59.83
7	3.00	468.90	28.00
8	3.50	539.80	13.13
9	4.00	575.40	6.19
10	4.50	589.10	2.33
11	5.00	592.80	0.62
12	5.50	595.70	0.49
13	6.00	598.20	0.42
14	6.50	600.30	0.35
15	7.00	602.30	0.33
16	7.50	603.90	0.26
17	8.00	605.10	0.20
18	8.50	606.30	0.20
19	9.00	607.50	0.20
20	9.50	608.80	0.21
21	10.00	609.60	0.13
22	10.50	610.50	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(610.5 - 11.75908) (1)/(1)	=	+5.987E+02	g/L Carbon
(610.5 - 11.75908) (1)/(1) (12)	=	+4.990E+01	Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3768

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 12/05/98

Time: 08:09:54

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.00	3.80	78.95
4	1.50	43.40	91.24
5	2.00	156.00	72.18
6	2.50	284.80	45.22
7	3.00	387.80	26.56
8	3.50	454.70	14.71
9	4.00	496.70	8.46
10	4.50	523.40	5.10
11	5.00	545.00	3.96
12	5.50	565.70	3.66
13	6.00	573.40	1.34
14	6.50	577.10	0.64
15	7.00	579.10	0.35
16	7.50	581.00	0.33
17	8.00	582.60	0.27
18	8.50	584.00	0.24
19	9.00	585.50	0.26
20	9.50	586.60	0.19
21	10.00	588.00	0.24
22	10.50	589.10	0.19

USER INPUT BLANK VALUE

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(589.1 - 39.47243) (1)/(1) = +5.496E+02 g/L Carbon
 (589.1 - 39.47243) (1)/(1) (12) = +4.580E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3769

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 12/05/98

Time: 08:38:41

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst :

JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	1.20	100.00
3	1.00	2.10	42.86
4	1.50	3.10	32.26
5	2.00	4.00	22.50
6	2.50	4.50	11.11
7	3.00	4.70	4.26
8	3.50	4.80	2.08
9	4.00	5.30	9.43
10	4.50	5.30	0.00
11	5.00	5.40	1.85
12	5.50	6.10	11.48
13	6.00	6.30	3.17
14	6.50	6.80	7.35
15	7.00	7.20	5.56
16	7.50	7.90	8.86
17	8.00	8.40	5.95
18	8.50	8.80	4.55
19	9.00	9.30	5.38
20	9.50	9.80	5.10
21	10.00	10.40	5.77
22	10.50	10.70	2.80

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(10.7 - 11.75891) (1) / (1) =

< 5.00 E-3 g/L Carbon

(10.7 - 11.75891) (1) / (1) (12) =

< 4.17 E-4 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3770

TOC

~~TIC~~ TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 12/05/98

Time: 08:54:06

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.40	100.00
3	1.01	4.60	69.57
4	1.51	8.40	45.24
5	2.00	11.90	29.41
6	2.50	14.20	16.20
7	3.00	16.40	13.41
8	3.50	17.90	8.38
9	4.00	19.20	6.77
10	4.50	20.40	5.88
11	5.00	21.40	4.67
12	5.50	22.50	4.89
13	6.00	23.70	5.06
14	6.50	24.90	4.82
15	7.00	25.90	3.86
16	7.50	27.10	4.43
17	8.00	28.10	3.56
18	8.50	29.10	3.44
19	9.00	30.10	3.32
20	9.50	31.00	2.90
21	10.00	31.60	1.90
22	10.50	32.30	2.17

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(32.3 - 11.75891) (1)/(1) = +2.05E+01 g/L Carbon
(32.3 - 11.75891) (1)/(1) (12) = +1.71E+00 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3771

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S97T002711

Date: 12/05/98

Time: 09:20:12

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	1.20	58.33
4	1.50	7.50	84.00
5	2.00	116.90	93.58
6	2.50	366.90	68.14
7	3.00	612.10	40.06
8	3.50	793.20	22.83
9	4.00	907.00	12.55
10	4.50	971.90	6.68
11	5.00	1006.80	3.47
12	5.50	1024.50	1.73
13	6.00	1033.60	0.88
14	6.50	1038.50	0.47
15	7.00	1041.10	0.25
16	7.50	1043.10	0.19
17	8.00	1044.80	0.16
18	8.50	1046.00	0.11
19	9.00	1047.10	0.11
20	9.50	1048.20	0.10
21	10.00	1049.30	0.10
22	10.50	1050.30	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

{ 1050.3 - 11.75884 } (1)/(1) = +1.0385E+03 g/L Carbon
 { 1050.3 - 11.75884 } (1)/(1) (12) = +8.6545E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3772

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002711

Date: 12/05/98

Time: 10:33:34

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.90	100.00
3	1.01	5.60	83.93
4	1.50	55.50	89.91
5	2.00	107.40	48.32
6	2.50	151.20	28.97
7	3.00	183.30	17.51
8	3.50	202.70	9.57
9	4.00	213.30	4.97
10	4.50	220.00	3.05
11	5.00	224.90	2.18
12	5.50	228.50	1.58
13	6.00	231.30	1.21
14	6.50	234.00	1.15
15	7.00	236.30	0.97
16	7.50	238.40	0.88
17	8.00	240.00	0.67
18	8.50	241.60	0.66
19	9.00	242.90	0.54
20	9.50	244.10	0.49
21	10.00	245.40	0.53
22	10.50	246.80	0.57

USER INPUT BLANK VALUE

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(246.8 - 39.47679) (1)/(1) = +2.073E+02 g/L Carbon
 (246.8 - 39.47679) (1)/(1) (12) = +1.728E+01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

JD SPELLMAN

22222

3773

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002711DUP

Date: 12/05/98

Time: 10:49:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	1.00	60.00
4	1.50	19.50	94.87
5	2.00	160.50	87.85
6	2.50	382.70	58.06
7	3.00	586.40	34.74
8	3.50	726.80	19.32
9	4.00	810.20	10.29
10	4.50	857.00	5.46
11	5.00	882.00	2.83
12	5.50	895.10	1.46
13	6.00	900.40	0.59
14	6.50	904.10	0.41
15	7.00	905.40	0.14
16	7.50	907.10	0.19
17	8.00	907.90	0.09
18	8.50	909.00	0.12
19	9.00	909.90	0.10
20	9.50	910.80	0.10
21	10.00	911.70	0.10
22	10.50	912.40	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(912.4 - 11.75781) (1) / (1)	=	+9.006E+02	g/L Carbon
(912.4 - 11.75781) (1) / (1) (12)	=	+7.505E+01	Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3774

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002711DUP

Date: 12/05/98

Time: 11:02:35

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	1.00	100.00
3	1.01	12.90	92.25
4	1.51	75.60	82.94
5	2.00	130.40	42.02
6	2.50	182.90	28.70
7	3.00	218.00	16.10
8	3.50	238.40	8.56
9	4.00	243.40	2.05
10	4.50	245.30	0.77
11	5.00	246.60	0.53
12	5.50	247.20	0.24
13	6.00	247.60	0.16
14	6.50	247.90	0.12
15	7.00	248.00	0.04
16	7.50	248.30	0.12
17	8.00	248.30	0.00
18	8.50	248.70	0.16
19	9.00	249.60	0.36
20	9.50	250.20	0.24
21	10.00	251.00	0.32
22	10.50	251.80	0.32

USER INPUT BLANK VALUE

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(251.8 - 39.47311) (1)/(1) = +2.123E+02 g/L Carbon
 (251.8 - 39.47311) (1)/(1)(12) = +1.769E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3775

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002711SPK

Date: 12/05/98

Time: 11:15:23

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	3.00	100.00
3	1.01	16.10	81.37
4	1.50	168.60	90.45
5	2.00	456.90	63.10
6	2.50	737.60	38.06
7	3.00	928.90	20.59
8	3.50	1035.50	10.29
9	4.00	1092.00	5.17
10	4.50	1121.70	2.65
11	5.00	1135.40	1.21
12	5.50	1141.40	0.53
13	6.00	1144.30	0.25
14	6.50	1146.20	0.17
15	7.00	1147.90	0.15
16	7.50	1149.10	0.10
17	8.00	1150.60	0.13
18	8.50	1151.60	0.09
19	9.00	1152.90	0.11
20	9.50	1153.80	0.08
21	10.00	1154.70	0.08
22	10.50	1155.60	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(1155.6 - 11.75795) (1)/(1) =

+1.1438E+03 g/L Carbon

(1155.6 - 11.75795) (1)/(1) (12) =

+9.5320E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3776

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002711SPK

Date: 12/05/98

Time: 11:30:09

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.01	83.50	99.04
4	1.51	301.50	72.31
5	2.00	505.40	40.34
6	2.50	686.10	26.34
7	3.00	782.30	12.30
8	3.50	831.00	5.86
9	4.00	858.20	3.17
10	4.50	874.20	1.83
11	5.00	884.50	1.16
12	5.50	891.80	0.82
13	6.00	897.50	0.64
14	6.50	901.80	0.48
15	7.00	905.40	0.40
16	7.50	908.30	0.32
17	8.00	911.20	0.32
18	8.50	913.20	0.22
19	9.03	915.50	0.25
20	9.53	916.90	0.15
21	10.03	918.90	0.22
22	10.53	920.30	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(920.3 - 39.60347) (1) / (1) = +8.807E+02 g/L Carbon
 (920.3 - 39.60347) (1) / (1) (12) = +7.339E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3777

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002840

Date: 12/05/98

Time: 11:49:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.30	46.15
4	1.50	3.00	56.67
5	2.00	9.10	67.03
6	2.50	30.10	69.77
7	3.00	67.40	55.34
8	3.50	114.00	40.88
9	4.00	156.20	27.02
10	4.50	185.10	15.61
11	5.00	203.10	8.86
12	5.50	214.50	5.31
13	6.00	221.90	3.33
14	6.50	227.30	2.38
15	7.00	233.30	2.57
16	7.50	237.20	1.64
17	8.00	239.40	0.92
18	8.50	241.10	0.71
19	9.00	242.20	0.45
20	9.50	243.20	0.41
21	10.00	244.20	0.41
22	10.50	245.40	0.49

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(245.4 - 11.75891) (1) / (1) = +2.336E+02 g/L Carbon
 (245.4 - 11.75891) (1) / (1) (12) = +1.947E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3778

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002840

Date: 12/05/98

Time: 12:01:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.00	24.70	96.76
4	1.50	66.70	62.97
5	2.00	115.10	42.05
6	2.50	166.60	30.91
7	3.00	202.80	17.85
8	3.50	224.20	9.55
9	4.00	236.90	5.36
10	4.50	245.00	3.31
11	5.00	250.00	2.00
12	5.50	254.20	1.65
13	6.00	257.70	1.36
14	6.50	261.40	1.42
15	7.00	263.50	0.80
16	7.50	265.70	0.83
17	8.00	267.70	0.75
18	8.50	269.30	0.59
19	9.00	270.90	0.59
20	9.50	272.30	0.51
21	10.00	273.50	0.44
22	10.50	274.80	0.47

USER INPUT BLANK VALUE

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(274.8 - 39.47289) (1) / (1) = +2.353E+02 g/L Carbon
 (274.8 - 39.47289) (1) / (1) (12) = +1.961E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002840DUP

Date: 12/05/98

Time: 12:19:19

Sample Size = 1 uL

Analyst : JD SPELLMAN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 1.12 ug/minute C

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.30	46.15
4	1.50	2.40	45.83
5	2.00	9.00	73.33
6	2.50	35.00	74.29
7	3.00	86.90	59.72
8	3.50	157.20	44.72
9	4.00	234.30	32.91
10	4.50	298.30	21.45
11	5.00	341.10	12.55
12	5.50	367.20	7.11
13	6.00	383.80	4.33
14	6.50	394.20	2.64
15	7.00	399.80	1.40
16	7.50	402.90	0.77
17	8.00	405.00	0.52
18	8.50	406.40	0.34
19	9.00	407.70	0.32
20	9.50	408.80	0.27
21	10.00	409.70	0.22
22	10.50	410.70	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(410.7 - 11.75781) (1)/(1) = +3.989E+02 g/L Carbon
 (410.7 - 11.75781) (1)/(1) (12) = +3.325E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3780

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002840SPK

Date: 12/05/98

Time: 12:57:49

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.12 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.00	3.30	75.76
4	1.50	104.40	96.84
5	2.00	427.80	75.60
6	2.50	776.80	44.93
7	3.00	1024.20	24.16
8	3.50	1167.40	12.27
9	4.00	1242.10	6.01
10	4.50	1285.30	3.36
11	5.00	1310.20	1.90
12	5.50	1325.70	1.17
13	6.00	1332.90	0.54
14	6.50	1336.00	0.23
15	7.00	1338.20	0.16
16	7.50	1340.30	0.16
17	8.00	1342.50	0.16
18	8.50	1344.60	0.16
19	9.00	1346.00	0.10
20	9.50	1347.50	0.11
21	10.00	1348.60	0.08
22	10.50	1349.50	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 11.75891 micrograms carbon

BLANK FACTOR = 11.75891 / 10.49902 = +1.1E+00 ug/min Carbon

SAMPLE RESULTS:

(1349.5 - 11.7598) (1)/(1) = +1.3377E+03 g/L Carbon
 (1349.5 - 11.7598) (1)/(1) (12) = +1.1148E+02 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3781

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002840SPK

Date: 12/05/98

Time: 16:04:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.76 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	17.40	95.98
4	1.50	89.50	80.56
5	2.00	210.40	57.46
6	2.50	337.00	37.57
7	3.00	414.90	18.78
8	3.50	459.00	9.61
9	4.00	485.60	5.48
10	4.50	504.60	3.77
11	5.00	518.10	2.61
12	5.50	525.40	1.39
13	6.00	529.60	0.79
14	6.50	532.50	0.54
15	7.00	535.00	0.47
16	7.50	537.30	0.43
17	8.00	539.10	0.33
18	8.50	540.70	0.30
19	9.00	542.20	0.28
20	9.50	543.40	0.22
21	10.00	544.60	0.22
22	10.50	545.70	0.20

USER INPUT BLANK VALUE:

BLANK VALUE = 39.47633 micrograms carbon

BLANK FACTOR = 39.47633 / 10.49902 = +3.8E+00 ug/min Carbon

SAMPLE RESULTS:

(545.7 - 39.4761) (1)/(1) = +5.062E+02 g/L Carbon
 (545.7 - 39.4761) (1)/(1) (12) = +4.219E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3782

LABCORE Completed Worklist Report for Worklist# 27372

Analyst: slh

Instrument: TOC01

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105, @TICTOC1, ReRun at <0.1g sample size. cjo

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	SOLID	6.02E+02	5.99E+2	99.502 % Recovery	
1 STD	0		@TICTOC1 TOC-02	SOLID	3.00E+03	2.82E+3	94.000 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	SOLID	1	4.50E+0	4.500 ug/g	
2 BLNK	0		@TICTOC1 TOC-02	SOLID	1	2.10E+0	2.100 ug/g	
3 SAMPLE	S98T002737	0	@TICTOC1 TIC-02	SOLID	N/A	2.40E+04	5.000 ug/g	
3 SAMPLE	S98T002737	0	@TICTOC1 TOC-02	SOLID	N/A	3.57E+3	40.000 ug/g	
4 DUP	S98T002737	0	@TICTOC1 TIC-02	SOLID	2.40E+4	1.46E+4	48.705 RPD	
4 DUP	S98T002737	0	@TICTOC1 TOC-02	SOLID	3.57E+3	3.18E+3	11.556 RPD	
5 TRIPL	S98T002737	0	@TICTOC1 TIC-02	SOLID	2.40E+4	1.32E+4	58.065 RPD	
5 TRIPL	S98T002737	0	@TICTOC1 TOC-02	SOLID	3.57E+3	<4.00E+1	RPD	
6 SPK	S98T002737	0	@TICTOC1 TIC-02	SOLID	1.00E+02	2.43E+01	24.300 % Recovery	
6 SPK	S98T002737	0	@TICTOC1 TOC-02	SOLID	1.00E+02	9.17E+01	91.700 % Recovery	
7 SAMPLE	S98T002914	0	@TICTOC1 TIC-02	SOLID	N/A	1.36E+04	5.000 ug/g	
7 SAMPLE	S98T002914	0	@TICTOC1 TOC-02	SOLID	N/A	7.84E+3	40.000 ug/g	
8 DUP	S98T002914	0	@TICTOC1 TIC-02	SOLID	1.36E+4	3.18E+4	80.176 RPD	
8 DUP	S98T002914	0	@TICTOC1 TOC-02	SOLID	7.84E+3	6.50E+3	18.689 RPD	
9 TRIPL	S98T002914	0	@TICTOC1 TIC-02	SOLID	1.36E+4	1.14E+4	17.600 RPD	
9 TRIPL	S98T002914	0	@TICTOC1 TOC-02	SOLID	7.84E+3	7.77E+3	0.897 RPD	
10 SPK	S98T002914	0	@TICTOC1 TIC-02	SOLID	1.00E+02	6.91E+01	69.100 % Recovery	
10 SPK	S98T002914	0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.19E+02	119.000 % Recovery	
11 SAMPLE	S98T003117	0	@TICTOC1 TIC-02	SOLID	N/A	7.84E+3	5.000 ug/g	
11 SAMPLE	S98T003117	0	@TICTOC1 TOC-02	SOLID	N/A	3.54E+3	40.000 ug/g	
12 DUP	S98T003117	0	@TICTOC1 TIC-02	SOLID	7.84E+3	7.74E+3	1.284 RPD	
12 DUP	S98T003117	0	@TICTOC1 TOC-02	SOLID	3.54E+3	3.62E+3	2.235 RPD	
13 SPK	S98T003117	0	@TICTOC1 TIC-02	SOLID	1.00E+02	9.91E+01	99.100 % Recovery	
13 SPK	S98T003117	0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.02E+02	102.000 % Recovery	

Final page for worklist# 27372

Analyst Signature _____ Date _____

Analyst Signature *C.D. Orin* Date *12/3/98*

Reviewer Signature *[Signature]* Date *12/4/98*

12/01/98 15:35

A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27372

Analyst: SLH Instrument: TOC01 Book# TK 25N12-DMethod: LA-342-100 Rev/Mod F-1

TOC 26N12-F

Worklist Comment: BY-105, @TICTOC1, cjo AND SX-102
TMS 12/1/98

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	SOLID		
2 BLNK			@TICTOC1	SOLID		
3 SAMPLE	S98T002914 0		@TICTOC1	SOLID	98000490	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002914 0		@TICTOC1	SOLID		
5 SPK	S98T002914 0		@TICTOC1	SOLID		
6 SAMPLE	S98T003117 0		@TICTOC1	SOLID	98000391	SX-102
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T003117 0		@TICTOC1	SOLID		
8 SPK	S98T003117 0		@TICTOC1	SOLID		

Final page for worklist # 27372

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

WL #27370 added to this WL
Batch 6027

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3784

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	10.9	18.3
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	µg of Carbon = C1-C2		
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006027			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
BLANK			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon	4.50E+00	2.10E+00
BY-105			

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>N/A</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3785

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	605	585.3
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000

Matrix
LIQUID

Batch Number
98006027

Rerun
0

Sample Prep
N/A

Sample #
STD

Instrument Code
TOC01

Prepared By
CJO

Chemist
MJL

Analyst
SLH

Date Complete
12/03/98

Analysis Date
12/02/98

Analysis Time
12:00 PM

Sample Point
BY-105

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
12:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.99E+02	2.82E+03
BY-105	Percent Standard Recovery	99.4	94.2

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>NA</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3786

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0313	0.0313
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	757.1	132.1
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code			
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98006027			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002737			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon/g	2.40E+04	3.57E+03
BY-105			

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0291	0.0291
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	430.3	113
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	Known µg of C from Original Sample	2.40E+4	3.57E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006027			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T002737			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon/g	1.46E+04	3.18E+03
BY-105			

$\mu\text{g of Carbon/g} = (C1 - C2) \times DF / SS$
 $\mu\text{g of Carbon/g for TIC} = 5 \text{ if } C1 < C2$
 $\mu\text{g of Carbon/g for TOC} = 40 \text{ if } C1 < C2$

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>MA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3788

WORKBOOK PAGE: TRIPL5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0316	0.0316
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	422.8	N/A
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	Known µg of C from Original Sample	2.40E+4	3.57E+3
@TICTOC1	Duplicate Result	1.46E+4	3.18E+3
Matrix			
SOLID			
Batch Number			
98006027			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002737			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date	NOTE: FOR TOC: The Reported Result is < 40.		
12/02/98			
Analysis Time	Method Detection Limit in ug/g		
12:00 PM			
Sample Point	µg of Carbon/g Note: <		
BY-105			

	TIC	TOC
	5	40

	TIC	TOC
	1.32E+04	4.00E+01

Data Entered By:

CJO

Date: 12/03/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: SPIKE6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0313	0.0313
Work List	Final Coulometer Reading in µg (C1)	757.1	132.1
27372	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0305	0.0305
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	811	404.2
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006027	µg C in baseline (BL)	6.4	20.4

Rerun
0Sample Prep
N/ASample #
S98T002737

$$\text{Percent Spike Recovery} = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100$$

Instrument Code
TOC01

$$QC \text{ Actual in } \mu\text{g/mL} = \text{Spike Value } (\mu\text{g/mL})$$

$$QC \text{ Found in } \mu\text{g/mL} = (\text{Percent Spike Recovery}) * (QC \text{ Actual}) / 100$$

Prepared By
CJOChemist
MJLAnalyst
SLHDate Complete
12/03/98Analysis Date
12/02/98

Analysis Time

12:00 PM

QC Actual in µg/mL

Sample Point

QC Found in µg/mL

BY-105

Percent Spike Recovery

TIC	TOC
6.02E+02	3.00E+03
1.46E+02	2.75E+03
24.3	91.7

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3790

WORKBOOK PAGE: SAM7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0447	0.0447
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	615.4	371
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006027			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002914			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon/g	1.36E+04	7.84E+03
BY-105			

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>N/A</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3791

WORKBOOK PAGE: DUP8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0476	0.0476
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	1519.2	329.9
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	Known µg of C from Original Sample	1.36E+4	7.84E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006027			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002914			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon/g	3.18E+04	6.50E+03
BY-105			

Data Entered By:

CJO

Date:

12/03/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: TRIPL9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0428	0.0428
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	494.1	352.9
27372	µg of Carbon from Baseline (C2)	6.4	20.4
Test Code	Known µg of C from Original Sample	1.36E+4	7.84E+3
@TICTOC1	Duplicate Result	3.18E+4	6.50E+3
Matrix			
SOLID			
Batch Number			
98006027			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002914			
Instrument Code			
TOC01			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time	Method Detection Limit in ug/g	5	40
12:00 PM			
Sample Point	µg of Carbon/g	1.14E+04	7.77E+03
BY-105			

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3793

WORKBOOK PAGE: SPIKE10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0447	0.0447
Worklist	Final Coulometer Reading in µg (C1)	615.4	371
27372	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0410	0.0410
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	772.9	699.6
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006027	µg C in baseline (BL)	6.4	20.4
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002914			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
TOC01	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/03/98			
Analysis Date			
12/02/98			
Analysis Time			
12:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	4.16E+02	3.58E+03
BY-105	Percent Spike Recovery	69.1	119.2

Data Entered By:	CJO	Date:	12/03/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3794

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/01/98

Time: 09:19:16

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.20	0.00
2	1.01	0.40	50.00
3	1.50	0.80	50.00
4	2.00	1.20	33.33
5	2.50	1.60	25.00
6	3.00	1.90	15.79
7	3.50	2.20	13.64
8	4.00	2.60	15.38
9	4.50	2.90	10.34
10	5.00	3.00	3.33
11	5.50	3.30	9.09
12	6.00	3.60	8.33
13	6.50	3.90	7.69
14	7.00	4.20	7.14
15	7.50	4.40	4.55
16	8.00	4.70	6.38
17	8.50	5.00	6.00
18	9.00	5.30	5.66
19	9.50	5.50	3.64
20	10.00	5.80	5.17
21	10.50	6.00	3.33
22	11.00	6.40	6.25

BLANK VALUE = 6.4 micrograms carbon

BLANK FACTOR = 6.4 / 10.99597 =

+5.8E-01

ug/min Carbon

Sample Run By:

SL HOOD

00002

3795

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/01/98

Time: 09:37:19

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.20	0.00
2	1.01	0.70	71.43
3	1.50	3.60	80.56
4	2.00	7.50	52.00
5	2.50	10.50	28.57
6	3.00	12.10	13.22
7	3.50	13.40	9.70
8	4.00	14.00	4.29
9	4.50	14.80	5.41
10	5.00	15.30	3.27
11	5.50	15.80	3.16
12	6.00	16.30	3.07
13	6.50	16.80	2.98
14	7.00	17.30	2.89
15	7.50	17.60	1.70
16	8.00	18.10	2.76
17	8.50	18.50	2.16
18	9.00	18.90	2.12
19	9.50	19.30	2.07
20	10.00	19.60	1.53
21	10.50	20.00	2.00
22	11.00	20.40	1.96

BLANK VALUE = 20.4 micrograms carbon

BLANK FACTOR = 20.4 / 10.99786 =

+1.85E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3796

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/01/98

Time: 10:29:59

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.50	0.00
2	1.00	1.30	61.54
3	1.50	2.90	55.17
4	2.00	3.90	25.64
5	2.50	4.50	13.33
6	3.00	5.10	11.76
7	3.50	5.50	7.27
8	4.00	5.90	6.78
9	4.50	6.20	4.84
10	5.00	6.60	6.06
11	5.50	7.00	5.71
12	6.00	7.30	4.11
13	6.50	7.70	5.19
14	7.00	8.10	4.94
15	7.50	8.40	3.57
16	8.00	8.70	3.45
17	8.50	9.10	4.40
18	9.00	9.50	4.21
19	9.50	9.80	3.06
20	10.00	10.20	3.92
21	10.50	10.50	2.86
22	11.00	10.90	3.67

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(10.9 - 6.378726) (1)/(1) = +4.52E+00 g/L Carbon
 (10.9 - 6.378726) (1)/(1) (12) = +3.77E-01 Molar Carbon

Sample Run By:

SL HOOD

00002

3797

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/01/98

Time: 10:42:46

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 1.85 ug/minute C

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.20	0.00
2	1.01	1.30	84.62
3	1.51	3.70	64.86
4	2.00	6.60	43.94
5	2.50	8.90	25.84
6	3.00	10.50	15.24
7	3.50	11.60	9.48
8	4.00	12.30	5.69
9	4.50	12.90	4.65
10	5.00	13.40	3.73
11	5.50	14.00	4.29
12	6.00	14.30	2.10
13	6.50	14.90	4.03
14	7.00	15.20	1.97
15	7.50	15.70	3.18
16	8.00	16.00	1.88
17	8.50	16.40	2.44
18	9.00	16.80	2.38
19	9.50	17.20	2.33
20	10.00	17.50	1.71
21	10.50	18.10	3.31
22	11.00	18.30	1.09

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(18.3 - 20.34447) (1) / (1) = < 5.00 E-3 g/L Carbon
 (18.3 - 20.34447) (1) / (1) (12) = < 4.17 E-4 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

SL HOOD

00002

3798

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/01/98

Time: 10:02:31

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	5.40	88.89
3	1.51	112.70	95.21
4	2.00	282.10	60.05
5	2.50	423.30	33.36
6	3.00	515.80	17.93
7	3.50	564.10	8.56
8	4.00	585.30	3.62
9	4.50	593.60	1.40
10	5.00	597.10	0.59
11	5.50	598.90	0.30
12	6.00	599.90	0.17
13	6.50	600.60	0.12
14	7.00	601.20	0.10
15	7.50	601.80	0.10
16	8.00	602.30	0.08
17	8.50	602.80	0.08
18	9.00	603.30	0.08
19	9.50	603.80	0.08
20	10.00	604.20	0.07
21	10.50	604.60	0.07
22	11.00	605.00	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(605 - 6.378265) (1)/(1000) = +5.986E-01 g/L Carbon
 (605 - 6.378265) (1)/(1000) (12) = +4.989E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3799 TO 3815.
 Sample Run By: SL HOOD 00002

1ml = 25N12-D

3799

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/01/98

Time: 10:17:33

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.40	71.43
3	1.51	25.40	94.49
4	2.00	118.80	78.62
5	2.50	293.70	59.55
6	3.00	452.70	35.12
7	3.50	530.60	14.68
8	4.00	559.80	5.22
9	4.50	571.00	1.96
10	5.00	575.40	0.76
11	5.50	577.50	0.36
12	6.00	578.90	0.24
13	6.50	580.00	0.19
14	7.00	580.70	0.12
15	7.50	581.40	0.12
16	8.00	582.10	0.12
17	8.50	582.60	0.09
18	9.00	583.20	0.10
19	9.50	583.80	0.10
20	10.00	584.30	0.09
21	10.50	584.70	0.07
22	11.00	585.30	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(585.3 - 20.34774) (1)/(200) = +2.825E+00 g/L Carbon
 (585.3 - 20.34774) (1)/(200) (12) = +2.354E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

.200 ml std 26012-F

3800

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002737

Date: 12/01/98

Time: 15:20:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	149.50	0.00
2	1.01	229.00	34.72
3	1.50	311.00	26.37
4	2.00	458.20	32.13
5	2.50	585.40	21.73
6	3.00	663.10	11.72
7	3.50	703.70	5.77
8	4.00	723.00	2.67
9	4.50	734.50	1.57
10	5.00	742.40	1.06
11	5.50	746.50	0.55
12	6.00	748.70	0.29
13	6.50	750.10	0.19
14	7.00	751.10	0.13
15	7.50	751.90	0.11
16	8.00	752.90	0.13
17	8.50	753.60	0.09
18	9.00	754.50	0.12
19	9.50	755.10	0.08
20	10.00	755.70	0.08
21	10.50	756.50	0.11
22	11.00	757.10	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(757.1 - 6.379327) (1)/(1) = +7.507E+02 g/L Carbon
 (757.1 - 6.379327) (1)/(1) (12) = +6.256E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

98 = .0313g

3801

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002737

Date: 12/01/98

Time: 15:33:57

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	3.40	79.41
3	1.51	34.70	90.20
4	2.00	71.00	51.13
5	2.50	96.60	26.50
6	3.00	110.20	12.34
7	3.50	116.10	5.08
8	4.00	119.10	2.52
9	4.50	121.10	1.65
10	5.00	122.70	1.30
11	5.50	123.90	0.97
12	6.00	124.90	0.80
13	6.50	125.90	0.79
14	7.00	126.80	0.71
15	7.50	127.50	0.55
16	8.00	128.20	0.55
17	8.50	129.00	0.62
18	9.00	129.60	0.46
19	9.50	130.30	0.54
20	10.00	130.90	0.46
21	10.50	131.50	0.46
22	11.00	132.10	0.45

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(132.1 - 20.34639) (1)/(1) = +1.118E+02 g/L Carbon
 (132.1 - 20.34639) (1)/(1) (12) = +9.313E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0313g

3802

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02737 DUP

Date: 12/01/98

Time: 15:47:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

```

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==
 1          0.51          0.60          0.00
 2          1.01          2.10          71.43
 3          1.51          69.20          96.97
 4          2.00          194.00         64.33
 5          2.50          295.70         34.39
 6          3.00          360.50         17.98
 7          3.50          395.30          8.80
 8          4.00          411.60          3.96
 9          4.50          418.10          1.55
10          5.00          420.90          0.67
11          5.50          422.40          0.36
12          6.00          423.50          0.26
13          6.50          424.40          0.21
14          7.00          425.30          0.21
15          7.50          426.00          0.16
16          8.00          426.80          0.19
17          8.50          427.30          0.12
18          9.00          427.90          0.14
19          9.50          428.50          0.14
20         10.00          429.20          0.16
21         10.50          429.70          0.12
22         11.00          430.30          0.14

```

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(430.3 - 6.377699) (1)/(1) = +4.239E+02 g/L Carbon
 (430.3 - 6.377699) (1)/(1) (12) = +3.533E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3803

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0.

Sample: S98T02737 DUP

Date: 12/01/98

Time: 15:59:45

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	3.10	80.65
3	1.51	28.70	89.20
4	2.01	59.80	52.01
5	2.50	82.00	27.07
6	3.00	93.80	12.58
7	3.50	98.90	5.16
8	4.00	101.60	2.66
9	4.50	103.20	1.55
10	5.00	104.40	1.15
11	5.50	105.50	1.04
12	6.00	106.50	0.94
13	6.50	107.40	0.84
14	7.00	107.90	0.46
15	7.50	108.70	0.74
16	8.00	109.30	0.55
17	8.50	110.00	0.64
18	9.00	110.70	0.63
19	9.50	111.20	0.45
20	10.00	111.80	0.54
21	10.50	112.50	0.62
22	11.00	113.00	0.44

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(113 - 20.34594) (1)/(1) = +9.265E+01 g/L Carbon
 (113 - 20.34594) (1)/(1) (12) = +7.721E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

55 = 0.0291g

3804

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02737 TRP

Date: 12/01/98

Time: 16:12:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	2.30	78.26
3	1.50	60.90	96.22
4	2.00	181.70	66.48
5	2.50	285.80	36.42
6	3.00	352.00	18.81
7	3.50	388.30	9.35
8	4.00	404.80	4.08
9	4.50	410.80	1.46
10	5.00	413.50	0.65
11	5.50	415.00	0.36
12	6.00	416.10	0.26
13	6.50	416.90	0.19
14	7.00	417.70	0.19
15	7.50	418.30	0.14
16	8.00	419.00	0.17
17	8.50	419.80	0.19
18	9.00	420.40	0.14
19	9.50	421.10	0.17
20	10.00	421.60	0.12
21	10.50	422.40	0.19
22	11.00	422.80	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(422.8 - 6.378761) (1) / (1) = +4.164E+02 g/L Carbon
 (422.8 - 6.378761) (1) / (1) (12) = +3.470E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = 03/leg

3805

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02737 SPK

Date: 12/01/98

Time: 16:25:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	3.20	81.25
3	1.50	116.30	97.25
4	2.00	368.00	68.40
5	2.50	575.40	36.04
6	3.00	692.90	16.96
7	3.50	754.50	8.16
8	4.00	784.20	3.79
9	4.50	795.50	1.42
10	5.00	799.80	0.54
11	5.50	802.00	0.27
12	6.00	803.30	0.16
13	6.50	804.40	0.14
14	7.00	805.30	0.11
15	7.50	806.10	0.10
16	8.00	806.80	0.09
17	8.50	807.70	0.11
18	9.00	808.40	0.09
19	9.50	809.00	0.07
20	10.00	809.60	0.07
21	10.50	810.30	0.09
22	11.00	811.00	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(811 - 6.379327) (1)/(1) = +8.046E+02 g/L Carbon
 (811 - 6.379327) (1)/(1) (12) = +6.705E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0305 + .500 ml spk 25N/2-D

3806

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02737SPK

Date: 12/01/98

Time: 16:38:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	Q.70	0.00
2	1.01	7.90	91.14
3	1.51	72.30	89.07
4	2.01	171.70	57.89
5	2.50	275.50	37.68
6	3.00	341.10	19.23
7	3.50	372.90	8.53
8	4.00	385.50	3.27
9	4.50	390.20	1.20
10	5.00	392.80	0.66
11	5.50	394.50	0.43
12	6.00	395.90	0.35
13	6.50	397.00	0.28
14	7.00	398.00	0.25
15	7.50	398.90	0.23
16	8.00	399.90	0.25
17	8.50	400.60	0.17
18	9.00	401.50	0.22
19	9.50	402.10	0.15
20	10.00	402.80	0.17
21	10.50	403.50	0.17
22	11.00	404.20	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(404.2 - 20.34447) (1)/(1) = +3.839E+02 g/L Carbon
 (404.2 - 20.34447) (1)/(1)(12) = +3.199E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = ,0305+, 100ml spk 26N12-F

3807

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002914

Date: 12/01/98

Time: 11:03:36

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .58 ug/minute C

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	1.00	70.00
3	1.50	16.10	93.79
4	2.00	110.00	85.36
5	2.50	264.50	58.41
6	3.00	403.20	34.40
7	3.50	499.60	19.30
8	4.00	552.10	9.51
9	4.50	577.30	4.37
10	5.00	589.00	1.99
11	5.50	595.40	1.07
12	6.00	599.60	0.70
13	6.50	603.00	0.56
14	7.00	605.50	0.41
15	7.50	607.60	0.35
16	8.00	609.50	0.31
17	8.50	611.00	0.25
18	9.00	612.30	0.21
19	9.50	613.20	0.15
20	10.00	614.20	0.16
21	10.50	614.80	0.10
22	11.00	615.40	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(615.4 - 6.378265) (1)/(1) = +6.090E+02 g/L Carbon
 (615.4 - 6.378265) (1)/(1) (12) = +5.075E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3808

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002914

Date: 12/01/98

Time: 11:15:58

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.80	0.00
2	1.01	3.70	78.38
3	1.51	62.00	94.03
4	2.00	147.40	57.94
5	2.50	218.90	32.66
6	3.00	264.90	17.37
7	3.50	292.40	9.40
8	4.00	310.40	5.80
9	4.50	323.40	4.02
10	5.00	333.00	2.88
11	5.50	340.50	2.20
12	6.00	346.30	1.67
13	6.50	351.10	1.37
14	7.00	355.00	1.10
15	7.50	358.30	0.92
16	8.00	361.20	0.80
17	8.50	363.40	0.61
18	9.00	365.50	0.57
19	9.50	367.10	0.44
20	10.00	368.60	0.41
21	10.50	369.90	0.35
22	11.00	371.00	0.30

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(371 - 20.34774) (1)/(1) = +3.507E+02 g/L Carbon
 (371 - 20.34774) (1)/(1) (12) = +2.922E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0447g

3809

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02914 DUP

Date: 12/01/98

Time: 11:28:35

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.00	1.70	58.82
3	1.50	49.40	96.56
4	2.00	326.60	84.87
5	2.50	781.90	58.23
6	3.00	1137.70	31.27
7	3.50	1332.60	14.63
8	4.00	1431.70	6.92
9	4.50	1478.50	3.17
10	5.00	1497.90	1.30
11	5.50	1505.30	0.49
12	6.00	1508.80	0.23
13	6.50	1510.90	0.14
14	7.00	1512.40	0.10
15	7.50	1512.60	0.08
16	8.00	1514.70	0.07
17	8.50	1515.40	0.05
18	9.00	1516.30	0.06
19	9.50	1517.00	0.05
20	10.00	1517.80	0.05
21	10.50	1518.50	0.05
22	11.00	1519.20	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(1519.2 - 6.378726) (1) / (1) = +1.5128E+03 g/L Carbon
 (1519.2 - 6.378726) (1) / (1) (12) = +1.2607E+02 Molar Carbon

Sample Run By:

SL HOOD

00002

35 = .0476g

3810

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02914 DUP

Date: 12/01/98

Time: 11:40:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

```

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==
  1          0.51          0.60          0.00
  2          1.01          7.40          91.89
  3          1.51          78.00          90.51
  4          2.00          156.90          50.29
  5          2.50          212.40          26.13
  6          3.00          246.20          13.73
  7          3.50          265.90          7.41
  8          4.00          278.30          4.46
  9          4.50          287.30          3.13
 10          5.00          294.70          2.51
 11          5.50          300.70          2.00
 12          6.00          305.50          1.57
 13          6.50          309.80          1.39
 14          7.00          313.20          1.09
 15          7.50          316.30          0.98
 16          8.00          319.00          0.85
 17          8.50          321.40          0.75
 18          9.00          323.50          0.65
 19          9.50          325.40          0.58
 20         10.00          327.20          0.55
 21         10.50          328.50          0.40
 22         11.00          329.90          0.42

```

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

```

( 329.9 - 20.34605 ) (1)/(1) = +3.096E+02 g/L Carbon
( 329.9 - 20.34605 ) (1)/(1) (12) = +2.580E+01 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

```

Sample Run By:

SL HOOD

00002

SS- .0476g

3811

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02914 TRP

Date: 12/01/98

Time: 11:52:57

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	1.10	0.00
2	1.01	2.50	56.00
3	1.50	27.50	90.91
4	2.00	110.60	75.14
5	2.50	226.00	51.06
6	3.00	335.90	32.72
7	3.50	411.70	18.41
8	4.00	453.10	9.14
9	4.50	472.50	4.11
10	5.00	480.60	1.69
11	5.50	484.20	0.74
12	6.00	486.10	0.39
13	6.50	487.50	0.29
14	7.00	488.60	0.23
15	7.50	489.50	0.18
16	8.00	490.20	0.14
17	8.50	491.10	0.18
18	9.00	491.70	0.12
19	9.50	492.40	0.14
20	10.00	493.00	0.12
21	10.50	493.60	0.12
22	11.00	494.10	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(494.1 - 6.378831) (1)/(1) = +4.877E+02 g/L Carbon
 (494.1 - 6.378831) (1)/(1) (12) = +4.064E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

35 = .0498g

3812

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02914 TRP

Date: 12/01/98

Time: 13:14:48

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	4.40	88.64
3	1.51	70.70	93.78
4	2.00	157.60	55.14
5	2.50	224.20	29.71
6	3.00	263.50	14.91
7	3.50	285.60	7.74
8	4.00	299.90	4.77
9	4.50	310.00	3.26
10	5.00	317.70	2.42
11	5.50	324.10	1.97
12	6.00	329.40	1.61
13	6.50	334.00	1.38
14	7.00	337.30	0.98
15	7.50	340.50	0.94
16	8.00	343.00	0.73
17	8.50	345.30	0.67
18	9.00	347.10	0.52
19	9.50	348.80	0.49
20	10.00	350.40	0.46
21	10.50	351.80	0.40
22	11.00	352.90	0.31

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(352.9 - 20.34932) (1)/(1) = +3.326E+02 g/L Carbon
 (352.9 - 20.34932) (1)/(1).(12) = +2.771E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0428g

3813

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTQC REV 2.0

Sample: S98T02914 SPK Date: 12/01/98 Time: 13:32:16

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = .58 ug/minute C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.50	73.33
3	1.51	55.70	97.31
4	2.00	230.50	75.84
5	2.50	431.30	46.56
6	3.00	583.70	26.11
7	3.50	678.10	13.92
8	4.00	728.10	6.87
9	4.50	750.40	2.97
10	5.00	758.90	1.12
11	5.50	762.50	0.47
12	6.00	764.40	0.25
13	6.50	765.90	0.20
14	7.00	766.90	0.13
15	7.50	767.80	0.12
16	8.00	768.70	0.12
17	8.50	769.60	0.12
18	9.00	770.20	0.08
19	9.50	771.00	0.10
20	10.00	771.80	0.10
21	10.50	772.30	0.06
22	11.00	772.90	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 6.378761 micrograms carbon

BLANK FACTOR = 6.378761 / 10.99786 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(772.9 - 6.378726) (1)/(1) = +7.665E+02 g/L Carbon
(772.9 - 6.378726) (1)/(1).(12) = +6.388E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

35 = .0410g + .5002g spk 25N12-D
3814

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02914 SPK

Date: 12/01/98

Time: 13:44:46

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.85 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.00	32.40	97.84
3	1.50	170.80	81.03
4	2.00	345.50	50.56
5	2.50	493.60	30.00
6	3.00	572.10	13.72
7	3.50	611.50	6.44
8	4.00	633.20	3.43
9	4.50	646.90	2.12
10	5.00	657.00	1.54
11	5.50	665.00	1.20
12	6.00	671.40	0.95
13	6.50	676.70	0.78
14	7.00	681.10	0.65
15	7.50	684.90	0.55
16	8.00	688.10	0.47
17	8.50	690.80	0.39
18	9.00	693.10	0.33
19	9.50	695.00	0.27
20	10.00	696.70	0.24
21	10.50	698.10	0.20
22	11.00	699.60	0.21

USER INPUT BLANK VALUE

BLANK VALUE = 20.34605 micrograms carbon

BLANK FACTOR = 20.34605 / 10.99786 = +1.9E+00 ug/min Carbon

SAMPLE RESULTS:

(699.6 - 20.34255) (1) / (1) = +6.793E+02 g/L Carbon
 (699.6 - 20.34255) (1) / (1) (12) = +5.660E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0410g + .100ml spk 26N12-F

3815

LABCORE Completed Worklist Report for Worklist# 27393

Analyst: slh

Instrument: TOC01

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY105 @TICTOC1 RERUN MF

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	@TICTOC1 TIC-02	SOLID	1	4.60E+0	4.600	ug/g
1 BLNK	0	@TICTOC1 TOC-02	SOLID	1	2.10E+0	2.100	ug/g
2 STD	0	@TICTOC1 TIC-02	SOLID	6.02E+02	5.82E+2	96.678	% Recovery
2 STD	0	@TICTOC1 TOC-02	SOLID	3.00E+03	2.81E+3	93.667	% Recovery
3 SAMPLE	S98T002980 0	@TICTOC1 TIC-02	SOLID	N/A	5.69E+03	5.000	ug/g
3 SAMPLE	S98T002980 0	@TICTOC1 TOC-02	SOLID	N/A	2.15E+03	40.000	ug/g
4 DUP	S98T002980 0	@TICTOC1 TIC-02	SOLID	5.69E+3	6.04E+3	5.968	RPD
4 DUP	S98T002980 0	@TICTOC1 TOC-02	SOLID	2.15E+3	2.24E+3	4.100	RPD
5 SPK	S98T002980 0	@TICTOC1 TIC-02	SOLID	1.00E+03	9.49E+01	94.900	% Recovery
5 SPK	S98T002980 0	@TICTOC1 TOC-02	SOLID	1.00E+02	7.08E+01	70.800	% Recovery
6 SAMPLE	S98T002981 0	@TICTOC1 TIC-02	SOLID	N/A	6.82E+03	5.000	ug/g
6 SAMPLE	S98T002981 0	@TICTOC1 TOC-02	SOLID	N/A	2.86E+03	40.000	ug/g
7 DUP	S98T002981 0	@TICTOC1 TIC-02	SOLID	6.82E+3	1.19E+4	54.274	RPD
7 DUP	S98T002981 0	@TICTOC1 TOC-02	SOLID	2.86E+3	3.40E+3	17.252	RPD
8 TRIPL	S98T002981 0	@TICTOC1 TIC-02	SOLID	6.82E+3	9.41E+3	31.916	RPD
8 TRIPL	S98T002981 0	@TICTOC1 TOC-02	SOLID	2.86E+3	3.51E+2	156.275	RPD
9 SPK	S98T002981 0	@TICTOC1 TIC-02	SOLID	1.00E+02	2.44E+02	244.000	% Recovery
9 SPK	S98T002981 0	@TICTOC1 TOC-02	SOLID	1.00E+02	1.01E+02	101.000	% Recovery
10 SAMPLE	S98T002983 0	@TICTOC1 TIC-02	SOLID	N/A	4.09E+03	5.000	ug/g
10 SAMPLE	S98T002983 0	@TICTOC1 TOC-02	SOLID	N/A	1.50E+03	40.000	ug/g
11 DUP	S98T002983 0	@TICTOC1 TIC-02	SOLID	4.09E+3	4.25E+3	3.837	RPD
11 DUP	S98T002983 0	@TICTOC1 TOC-02	SOLID	1.50E+3	1.91E+3	24.047	RPD
12 TRIPL	S98T002983 0	@TICTOC1 TIC-02	SOLID	4.09E+3	3.97E+3	2.978	RPD
12 TRIPL	S98T002983 0	@TICTOC1 TOC-02	SOLID	1.50E+3	1.76E+3	15.951	RPD

Final page for worklist# 27393

Analyst Signature _____ Date _____

Analyst Signature *Yvonne Frey* Date *12-4-98*

Reviewer Signature *[Signature]* Date *12/8/98*

12/02/98 13:28
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27393

Analyst: JH Instrument: TOC01 _____ Book# TIC 25N12-D
 Method: LA-342-100 Rev/Mod F-1 TOC 26M12-F

Worklist Comment: BY105 @TICTOC1 RERUN MF

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	SOLID		
2 STD			@TICTOC1	SOLID		
3 SAMPLE	S98T002980 0		@TICTOC1	SOLID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002980 0		@TICTOC1	SOLID		
5 SPK	S98T002980 0		@TICTOC1	SOLID		
6 SAMPLE	S98T002981 0		@TICTOC1	SOLID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
7 DUP	S98T002981 0		@TICTOC1	SOLID		
8 SPK	S98T002981 0		@TICTOC1	SOLID		

Final page for worklist # 27393

Jordan Howard Beatty
 Analyst Signature Date 12-3-98

Yvonne Jones 12-4-98
 Analyst Signature Date

Data Entry Comments:

Sample Dup + Trip for S98T002983 added
to this WL

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3817

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	ug of Carbon in Sample (C1)	11	26.1
27393	ug of Carbon from Baseline (C2)	6.4	24
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	ug of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	ug of Carbon	4.60E+00	2.10E+00
BY 105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3818

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	588.8	585
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000
Matrix			
LIQUID			
Batch Number			
98006014			
Rerun			
0	QC Actual in µg/mL = Standard Value (µg/mL)		
Sample/Prep	QC Found in µg/mL = (C1 - C2) * DF / SS		
N/A	QC Found in µg/mL for TIC = 5 if C1 < C2		
Sample #	QC Found in µg/mL for TOC = 40 if C1 < C2		
STD			
Instrument Code	% Recovery = QC Found / QC Actual * 100		
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
05:30 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.82E+02	2.81E+03
BY. 105	Percent Standard Recovery	96.7	93.5

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>MF</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3819

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0652	0.0652
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	377.2	164.3
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002980			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g	5.69E+03	2.15E+03
BY 105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3820

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0706	0.0706
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	432.9	182.4
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Known µg of C from Original Sample	5.69E+3	2.15E+3
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002980			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:30 PM			
Sample Point	µg of Carbon/g	6.04E+03	2.24E+03
BY 105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3821

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0652	0.0652
Work List	Final Coulometer Reading in µg (C1)	377.2	164.3
27393	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0682	0.0682
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	679.8	383.2
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006014	µg C in baseline (BL)		

Rerun
0

Sample Prep
N/A

Sample #
S98T002980

Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100

Instrument Code
TOC01

QC Actual in µg/mL = Spike Value (µg/mL)

QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100

Prepared By
MF

Chemist
MJL

Analyst
SLH

Date Complete
12/04/98

Analysis Date
12/03/98

Analysis Time

05:30 PM

QC Actual in µg/mL

Sample Point

QC Found in µg/mL

BY 105

Percent Spike Recovery

TIC	TOC
6.02E+02	3.00E+03
5.70E+02	2.11E+03
94.8	70.4

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>NA</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3822

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Size in g (SS)	TIC	TOC
SAMPLE	Dilution Factor (DF)	0.0795	0.0795
Work List	μg of Carbon in Sample (C1)	1	1
27393	μg of Carbon from Baseline (C2)	548.9	251.6
Test Code		6.4	24
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	μg of Carbon/g = (C1-C2) * DF / SS		
0	μg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	μg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002981			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:30 PM		5	40
Sample Point	μg of Carbon/g	6.82E+03	2.86E+03
BY 105			

Data Entered By:

MF

Date: 12/04/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0797	0.0797
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	952.9	295
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Known µg of C from Original Sample	6.82E+3	2.86E+3
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002981			
Instrument Code			
TOC01			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:30 PM		5	40
Sample Point	µg of Carbon/g	1.19E+04	3.40E+03
BY 105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>N/A</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3824

WORKBOOK PAGE: TRIPL8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0855	0.8550
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	810.8	324.3
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Known µg of C from Original Sample	6.82E+3	2.86E+3
@TICTOC1	Duplicate Result	1.19E+4	3.40E+3
Matrix	µg of Carbon/g = (C1-C2) * DF / SS µg of Carbon/g for TIC = 5 if C1 < C2 µg of Carbon/g for TOC = 40 if C1 < C2		
SOLID			
Batch Number			
98006014			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T002981			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
03:00 PM		5	40
Sample Point	µg of Carbon/g	9.41E+03	3.51E+02
BY-105			

Data Entered By:

MF

Date: 12/04/98

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: SPIKE9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0795	0.0795
Work List	Final Coulometer Reading in µg (C1)	548.9	251.6
27393	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0793	0.0793
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	1281.3	554.9
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006014	µg C in baseline (BL)	6.4	24
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T002981			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time			
03:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	1.47E+03	3.04E+03
BY-105	Percent Spike Recovery	243.8	101.3

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>N/A</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3826

WORKBOOK PAGE: SAM10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0601	0.0601
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	252.5	114.4
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002983			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	5	40
03:00 PM			
Sample Point	µg of Carbon/g	4.09E+03	1.50E+03
BY-105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3827

WORKBOOK PAGE: DUP11

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0698	0.0698
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	303.1	157.2
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Known µg of C from Original Sample	4.20E+6	1.50E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006014			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002983			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	5	40
03:00 PM			
Sample Point	µg of Carbon/g	4.25E+03	1.91E+03
BY-105			

Data Entered By:	MF	Date:	12/04/98
Signature of Chemist:	NA	Date:	

SAMPLE WB1 REV 1.0

342100ML

3828

WORKBOOK PAGE: TRIPL12

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0634	0.0634
TRIPL	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	258.3	135.9
27393	µg of Carbon from Baseline (C2)	6.4	24
Test Code	Known µg of C from Original Sample	4.20E+6	1.50E+3
@TICTOC1	Duplicate Result	4.25E+3	1.91E+3
Matrix	µg of Carbon/g = (C1-C2) * DF / SS µg of Carbon/g for TIC = 5 if C1 < C2 µg of Carbon/g for TOC = 40 if C1 < C2		
SOLID			
Batch Number			
98006014			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T002983			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/04/98			
Analysis Date			
12/03/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
03:00 PM		5	40
Sample Point	µg of Carbon/g	3.97E+03	1.76E+03
BY-105			

Data Entered By:

MF

Date: 12/04/98

Signature of Chemist:

NA

Date:

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/03/98

Time: 10:10:21

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.00	68.10	98.97
3	1.50	226.80	69.97
4	2.00	384.30	40.98
5	2.50	486.90	21.07
6	3.00	542.50	10.25
7	3.50	566.80	4.29
8	4.00	576.20	1.63
9	4.50	579.80	0.62
10	5.00	581.50	0.29
11	5.50	582.50	0.17
12	6.00	583.20	0.12
13	6.50	583.90	0.12
14	7.00	584.50	0.10
15	7.50	585.10	0.10
16	8.00	585.60	0.09
17	8.50	586.20	0.10
18	9.00	586.70	0.09
19	9.50	587.30	0.10
20	10.00	587.70	0.07
21	10.50	588.30	0.10
22	11.00	588.80	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(588.8 - 6.377203) (1)/(1000) = +5.824E-01 g/L Carbon
 (588.8 - 6.377203) (1)/(1000) (12) = +4.854E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3830 TO 3855.Sample Run By: SL Hood

SL HOOD

00002

80 = 1ml = 25N12-D

3830

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/03/98

Time: 10:24:59

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.00	1.40	64.29
3	1.50	8.30	83.13
4	2.00	39.20	78.83
5	2.50	120.90	67.58
6	3.00	273.90	55.86
7	3.50	432.10	36.61
8	4.00	516.40	16.32
9	4.50	548.30	5.82
10	5.00	559.90	2.07
11	5.50	564.80	0.87
12	6.00	568.90	0.72
13	6.50	571.70	0.49
14	7.00	573.40	0.30
15	7.50	575.20	0.31
16	8.00	577.60	0.42
17	8.50	579.00	0.24
18	9.00	580.10	0.19
19	9.50	581.40	0.22
20	10.00	582.90	0.26
21	10.50	584.10	0.21
22	11.00	585.00	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(585 - 23.97335) (1)/(200) = +2.805E+00 g/L Carbon
 (585 - 23.97335) (1)/(200) (12) = +2.338E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

20 = 200 mL = 200/12-F
3831

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/03/98

Time: 09:36:38

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	0.20	0.00
2	1.01	0.50	60.00
3	1.51	0.90	44.44
4	2.01	1.40	35.71
5	2.51	1.80	22.22
6	3.01	2.10	14.29
7	3.51	2.40	12.50
8	4.00	2.70	11.11
9	4.50	2.90	6.90
10	5.00	3.20	9.37
11	5.50	3.50	8.57
12	6.00	3.70	5.41
13	6.50	4.00	7.50
14	7.00	4.20	4.76
15	7.50	4.50	6.67
16	8.00	4.80	6.25
17	8.50	5.00	4.00
18	9.00	5.40	7.41
19	9.50	5.60	3.57
20	10.00	5.90	5.08
21	10.50	6.20	4.84
22	11.00	6.40	3.13

BLANK VALUE = 6.4 micrograms carbon
 BLANK FACTOR = 6.4 / 10.99963 =

+5.8E-01 ug/min Carbon

Sample Run By:

SL HOOD

00002

3832

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/03/98

Time: 09:54:29

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.20	0.00
2	1.01	0.80	75.00
3	1.50	4.20	80.95
4	2.00	9.10	53.85
5	2.50	12.70	28.35
6	3.00	14.80	14.19
7	3.50	16.10	8.07
8	4.00	17.10	5.85
9	4.50	17.70	3.39
10	5.00	18.30	3.28
11	5.50	18.90	3.17
12	6.00	19.60	3.57
13	6.50	20.10	2.49
14	7.00	20.60	2.43
15	7.50	21.10	2.37
16	8.00	21.60	2.31
17	8.50	22.00	1.82
18	9.00	22.40	1.79
19	9.50	22.80	1.75
20	10.00	23.20	1.72
21	10.50	23.60	1.69
22	11.00	24.00	1.67

BLANK VALUE = 24 micrograms carbon

BLANK FACTOR = 24 / 10.99884 =

+2.18E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3833

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/03/98

Time: 12:52:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.90	55.56
3	1.50	1.60	43.75
4	2.00	2.20	27.27
5	2.50	2.80	21.43
6	3.00	3.40	17.65
7	3.50	3.90	12.82
8	4.00	4.30	9.30
9	4.50	4.90	12.24
10	5.00	5.40	9.26
11	5.50	5.80	6.90
12	6.00	6.30	7.94
13	6.50	6.70	5.97
14	7.00	7.20	6.94
15	7.50	7.70	6.49
16	8.00	8.20	6.10
17	8.50	8.60	4.65
18	9.00	9.10	5.49
19	9.50	9.60	5.21
20	10.00	10.10	4.95
21	10.50	10.60	4.72
22	11.00	11.00	3.64

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(11 - 6.379327) (1)/(1) =

+4.62E+00 g/L Carbon

(11 - 6.379327) (1)/(1) (12) =

+3.85E-01 Molar Carbon

Sample Run By:

SL HOOD

00002

3834

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/03/98

Time: 13:05:19

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.00	50.00
3	1.51	3.90	74.36
4	2.00	9.10	57.14
5	2.50	13.20	31.06
6	3.01	15.80	16.46
7	3.51	17.20	8.14
8	4.01	18.30	6.01
9	4.50	19.10	4.19
10	5.00	19.80	3.54
11	5.50	20.30	2.46
12	6.00	20.90	2.87
13	6.50	21.50	2.79
14	7.00	22.00	2.27
15	7.50	22.50	2.22
16	8.00	23.10	2.60
17	8.50	23.70	2.53
18	9.00	24.20	2.07
19	9.50	24.80	2.42
20	10.00	25.30	1.98
21	10.50	25.70	1.56
22	11.00	26.10	1.53

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(26.1 - 23.97561) (1)/(1) = +2.12E+00 g/L Carbon
 (26.1 - 23.97561) (1)/(1) (12) = +1.77E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3835

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002980

Date: 12/03/98

Time: 13:17:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.30	61.54
3	1.51	37.70	96.55
4	2.00	133.20	71.70
5	2.50	231.50	42.46
6	3.00	299.60	22.73
7	3.50	337.80	11.31
8	4.00	355.60	5.01
9	4.50	362.80	1.98
10	5.00	366.00	0.87
11	5.50	367.70	0.46
12	6.00	368.80	0.30
13	6.50	369.70	0.24
14	7.00	370.60	0.24
15	7.50	371.30	0.19
16	8.00	373.50	0.59
17	8.50	374.00	0.13
18	9.00	374.60	0.16
19	9.50	375.30	0.19
20	10.00	375.90	0.16
21	10.50	376.50	0.16
22	11.00	377.20	0.19

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(377.2 - 6.378831) (1) / (1)	=	+3.708E+02	g/L Carbon
(377.2 - 6.378831) (1) / (1) (12)	=	+3.090E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

3836

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002980

Date: 12/03/98

Time: 13:29:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.70	0.00
2	1.01	6.70	89.55
3	1.50	53.00	87.36
4	2.00	97.00	45.36
5	2.50	124.70	22.21
6	3.00	138.90	10.22
7	3.50	145.80	4.73
8	4.00	149.70	2.61
9	4.50	152.10	1.58
10	5.00	153.90	1.17
11	5.50	155.40	0.97
12	6.00	156.50	0.70
13	6.50	157.50	0.63
14	7.00	158.50	0.63
15	7.50	159.50	0.63
16	8.00	160.40	0.56
17	8.50	161.00	0.37
18	9.00	161.70	0.43
19	9.50	162.40	0.43
20	10.00	163.10	0.43
21	10.50	163.70	0.37
22	11.00	164.30	0.37

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(164.3 - 23.97162) (1)/(1) = +1.403E+02 g/L Carbon
 (164.3 - 23.97162) (1)/(1)(12) = +1.169E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0652g

3837

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV. 2.0

Sample: S98T02980 DUP

Date: 12/03/98

Time: 13:42:18

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	2.30	78.26
3	1.51	58.60	96.08
4	2.01	180.00	67.44
5	2.50	286.60	37.19
6	3.00	355.80	19.45
7	3.50	390.30	8.84
8	4.00	406.60	4.01
9	4.50	414.00	1.79
10	5.00	417.80	0.91
11	5.50	420.40	0.62
12	6.00	422.60	0.52
13	6.50	424.40	0.42
14	7.00	425.80	0.33
15	7.50	427.10	0.30
16	8.00	428.10	0.23
17	8.50	428.90	0.19
18	9.00	429.80	0.21
19	9.50	430.60	0.19
20	10.00	431.30	0.16
21	10.50	432.10	0.19
22	11.00	432.90	0.18

USER INPUT BLANK VALUE.

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(432.9 - 6.378265) (1)/(1) = +4.265E+02 g/L Carbon
 (432.9 - 6.378265) (1)/(1) (12) = +3.554E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0706 g

3838

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02980 DUP

Date: 12/03/98

Time: 13:55:02

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.80	0.00
2	1.01	3.10	74.19
3	1.51	32.30	90.40
4	2.00	79.20	59.22
5	2.50	119.10	33.50
6	3.00	143.10	16.77
7	3.50	155.30	7.86
8	4.00	162.20	4.25
9	4.50	166.20	2.41
10	5.00	168.90	1.60
11	5.50	170.90	1.17
12	6.00	172.50	0.93
13	6.50	174.00	0.86
14	7.00	175.30	0.74
15	7.50	176.40	0.62
16	8.00	177.40	0.56
17	8.50	178.30	0.50
18	9.00	179.20	0.50
19	9.50	180.20	0.55
20	10.00	180.90	0.39
21	10.50	181.70	0.44
22	11.00	182.40	0.38

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(182.4 - 23.97162) (1)/(1) = +1.584E+02 g/L Carbon
 (182.4 - 23.97162) (1)/(1) (12) = +1.320E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS- 0704g

3839

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02980 SPK

Date: 12/03/98

Time: 14:07:41

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.90	63.16
3	1.50	77.00	97.53
4	2.00	278.00	72.30
5	2.50	459.30	39.47
6	3.00	571.60	19.65
7	3.50	627.90	8.97
8	4.00	652.60	3.78
9	4.50	663.00	1.57
10	5.00	667.60	0.69
11	5.50	670.00	0.36
12	6.00	671.50	0.22
13	6.50	672.70	0.18
14	7.00	673.60	0.13
15	7.50	674.50	0.13
16	8.00	675.30	0.12
17	8.50	676.20	0.13
18	9.00	676.90	0.10
19	9.50	677.70	0.12
20	10.00	678.40	0.10
21	10.50	679.10	0.10
22	11.00	679.80	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(679.8 - 6.37823) (1)/(1) = +6.734E+02 g/L Carbon
 (679.8 - 6.37823) (1)/(1) (12) = +5.612E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0682g + .500 ml spk 25N/2-D

3840

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02980 SPK

Date: 12/03/98

Time: 14:22:30

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	4.40	84.09
3	1.50	28.80	84.72
4	2.00	100.30	71.29
5	2.50	216.80	53.74
6	3.00	302.30	28.28
7	3.50	342.10	11.63
8	4.00	358.80	4.65
9	4.50	366.00	1.97
10	5.00	369.50	0.95
11	5.50	371.80	0.62
12	6.00	373.50	0.46
13	6.50	374.90	0.37
14	7.00	376.00	0.29
15	7.50	377.20	0.32
16	8.00	378.20	0.26
17	8.50	379.20	0.26
18	9.00	380.10	0.24
19	9.50	380.90	0.21
20	10.00	381.70	0.21
21	10.50	382.50	0.21
22	11.00	383.20	0.18

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(383.2 - 23.97348) (1)/(1) = +3.592E+02 g/L Carbon
 (383.2 - 23.97348) (1)/(1) (12) = +2.994E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0682g + .100ml spk 26N12-F

3841

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002981

Date: 12/03/98

Time: 14:35:23

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.40	50.00
3	1.50	11.00	87.27
4	2.00	100.00	89.00
5	2.50	243.10	58.86
6	3.00	368.50	34.03
7	3.50	451.00	18.29
8	4.00	496.70	9.20
9	4.50	520.70	4.61
10	5.00	531.80	2.09
11	5.50	536.70	0.91
12	6.00	539.40	0.50
13	6.50	541.10	0.31
14	7.00	542.30	0.22
15	7.50	543.30	0.18
16	8.00	544.30	0.18
17	8.50	545.20	0.17
18	9.00	545.90	0.13
19	9.50	546.70	0.15
20	10.00	547.40	0.13
21	10.50	548.10	0.13
22	11.00	548.90	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(548.9 - 6.378265) (1) / (1) = +5.425E+02 g/L Carbon
 (548.9 - 6.378265) (1) / (1) (12) = +4.521E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS- .0795g

3842

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002981

Date: 12/03/98

Time: 14:47:39

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 2.18 ug/minute C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.80	0.00
2	1.01	2.60	69.23
3	1.50	52.00	95.00
4	2.00	124.20	58.13
5	2.50	176.60	29.67
6	3.00	207.10	14.73
7	3.50	222.90	7.09
8	4.00	230.50	3.30
9	4.50	234.90	1.87
10	5.00	237.70	1.18
11	5.50	239.80	0.88
12	6.00	241.60	0.75
13	6.50	242.90	0.54
14	7.00	244.10	0.49
15	7.50	245.30	0.49
16	8.00	246.40	0.45
17	8.50	247.60	0.48
18	9.00	248.40	0.32
19	9.50	249.20	0.32
20	10.00	250.10	0.36
21	10.50	250.90	0.32
22	11.00	251.60	0.28

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(251.6 - 23.97122) (1)/(1) = +2.276E+02 g/L Carbon
 (251.6 - 23.97122) (1)/(1) (12) = +1.897E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = 07953

3843

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 DUP

Date: 12/03/98

Time: 15:00:58

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	2.00	65.00
3	1.51	101.60	98.03
4	2.00	376.50	73.01
5	2.50	619.70	39.24
6	3.00	776.70	20.21
7	3.50	862.50	9.95
8	4.00	905.00	4.70
9	4.50	925.90	2.26
10	5.00	935.50	1.03
11	5.50	939.80	0.46
12	6.00	942.10	0.24
13	6.50	943.80	0.18
14	7.00	945.20	0.15
15	7.50	946.50	0.14
16	8.00	947.60	0.12
17	8.50	948.70	0.12
18	9.00	949.50	0.08
19	9.50	950.40	0.09
20	10.00	951.30	0.09
21	10.50	952.10	0.08
22	11.00	952.90	0.08

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(952.9 - 6.378761) (1)/(1) = +9.465E+02 g/L Carbon
 (952.9 - 6.378761) (1)/(1) (12) = +7.888E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3844

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 DUP

Date: 12/03/98

Time: 15:14:40

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	2.70	77.78
3	1.51	36.50	92.60
4	2.00	107.30	65.98
5	2.50	175.90	39.00
6	3.00	222.00	20.77
7	3.50	248.60	10.70
8	4.00	263.70	5.73
9	4.50	271.50	2.87
10	5.00	276.20	1.70
11	5.50	279.50	1.18
12	6.00	282.20	0.96
13	6.50	284.20	0.70
14	7.00	286.10	0.66
15	7.50	287.50	0.49
16	8.00	288.90	0.48
17	8.50	290.10	0.41
18	9.00	291.20	0.38
19	9.50	292.30	0.38
20	10.00	293.30	0.34
21	10.50	294.10	0.27
22	11.00	295.00	0.31

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(295 - 23.97747) (1)/(1) = +2.710E+02 g/L Carbon
 (295 - 23.97747) (1)/(1) (12) = +2.259E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .07973

3845

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 TRP

Date: 12/03/98

Time: 15:27:45

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.60	0.00
2	1.01	1.50	60.00
3	1.50	59.70	97.49
4	2.00	275.10	78.30
5	2.50	492.30	44.12
6	3.00	641.30	23.23
7	3.50	721.70	11.14
8	4.00	761.70	5.25
9	4.50	782.80	2.70
10	5.00	792.60	1.24
11	5.50	796.70	0.51
12	6.00	799.10	0.30
13	6.50	800.80	0.21
14	7.00	802.40	0.20
15	7.50	803.60	0.15
16	8.00	804.80	0.15
17	8.50	805.90	0.14
18	9.00	807.00	0.14
19	9.50	808.00	0.12
20	10.00	808.90	0.11
21	10.50	809.80	0.11
22	11.00	810.80	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(810.8 - 6.378265) (1) / (1) =

+8.044E+02 g/L Carbon

(810.8 - 6.378265) (1) / (1) (12) =

+6.704E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3846

SS = .0855g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 TRP

Date: 12/03/98

Time: 15:40:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.80	0.00
2	1.00	8.70	90.80
3	1.50	126.90	93.14
4	2.00	229.50	44.71
5	2.50	270.50	15.16
6	3.00	291.30	7.14
7	3.50	301.60	3.42
8	4.00	306.40	1.57
9	4.50	309.30	0.94
10	5.00	311.70	0.77
11	5.50	313.50	0.57
12	6.00	314.90	0.44
13	6.50	316.20	0.41
14	7.00	317.40	0.38
15	7.50	318.40	0.31
16	8.00	319.50	0.34
17	8.50	320.40	0.28
18	9.00	321.10	0.22
19	9.50	322.00	0.28
20	10.00	322.80	0.25
21	10.50	323.60	0.25
22	11.00	324.30	0.22

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(324.3 - 23.97122) (1)/(1) = +3.003E+02 g/L Carbon
 (324.3 - 23.97122) (1)/(1) (12) = +2.503E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0855g

3847

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 SPK

Date: 12/03/98

Time: 15:54:15

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.40	0.00
2	1.00	2.00	80.00
3	1.50	155.50	98.71
4	2.00	561.30	72.30
5	2.50	897.20	37.44
6	3.00	1092.50	17.88
7	3.50	1185.50	7.84
8	4.00	1227.30	3.41
9	4.50	1249.70	1.79
10	5.00	1261.50	0.94
11	5.50	1265.70	0.33
12	6.00	1268.20	0.20
13	6.50	1270.30	0.17
14	7.00	1271.90	0.13
15	7.50	1273.30	0.11
16	8.00	1274.70	0.11
17	8.50	1275.80	0.09
18	9.00	1277.00	0.09
19	9.50	1278.00	0.08
20	10.00	1279.10	0.09
21	10.50	1280.20	0.09
22	11.00	1281.30	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(1281.3 - 6.378761) (1) / (1) =

+1.2749E+03 g/L Carbon

(1281.3 - 6.378761) (1) / (1) (12) =

+1.0624E+02 Molar Carbon

Sample Run By:

SL HOOD

00002

38 = .07933 + .500 ml spk 25N/2-D

3848

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02981 SPK

Date: 12/03/98

Time: 16:06:35

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference
1	0.51	0.70	0.00
2	1.01	9.00	92.22
3	1.50	98.70	90.88
4	2.00	240.80	59.01
5	2.50	379.90	36.61
6	3.00	456.20	16.73
7	3.50	493.30	7.52
8	4.00	515.10	4.23
9	4.50	528.60	2.55
10	5.00	534.30	1.07
11	5.50	537.50	0.60
12	6.00	540.50	0.56
13	6.50	542.70	0.41
14	7.00	544.50	0.33
15	7.50	546.20	0.31
16	8.00	547.60	0.26
17	8.50	549.10	0.27
18	9.00	550.40	0.24
19	9.50	551.50	0.20
20	10.00	552.70	0.22
21	10.50	553.80	0.20
22	11.00	554.90	0.20

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(554.9 - 23.97521) (1)/(1) = +5.309E+02 g/L Carbon
 (554.9 - 23.97521) (1)/(1) (12) = +4.424E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0793g + .100 ml spk 26N12-F

3849

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002983

Date: 12/03/98

Time: 16:18:49

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.01	1.50	60.00
3	1.50	11.60	87.07
4	2.00	42.40	72.64
5	2.50	107.60	60.59
6	3.00	167.70	35.84
7	3.50	202.70	17.27
8	4.00	223.60	9.35
9	4.50	236.00	5.25
10	5.00	239.50	1.46
11	5.50	241.30	0.75
12	6.00	242.90	0.66
13	6.50	244.00	0.45
14	7.00	245.00	0.41
15	7.50	246.00	0.41
16	8.00	247.00	0.40
17	8.50	247.90	0.36
18	9.00	248.90	0.40
19	9.50	249.70	0.32
20	10.00	250.60	0.36
21	10.50	251.40	0.32
22	11.00	252.50	0.44

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(252.5 - 6.379327) (1)/(1)	=	+2.461E+02	g/L Carbon
(252.5 - 6.379327) (1)/(1) (12)	=	+2.051E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

55-0601g

3850

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002983

Date: 12/03/98

Time: 16:31:15

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.00	0.00
2	1.01	3.30	100.00
3	1.51	23.70	86.08
4	2.00	50.50	53.07
5	2.50	71.60	29.47
6	3.00	86.20	16.94
7	3.50	93.50	7.81
8	4.00	96.20	2.81
9	4.50	98.50	2.34
10	5.00	100.40	1.89
11	5.50	101.90	1.47
12	6.00	103.60	1.64
13	6.50	104.90	1.24
14	7.00	106.10	1.13
15	7.50	107.40	1.21
16	8.00	108.50	1.01
17	8.50	109.30	0.73
18	9.00	110.40	1.00
19	9.50	111.60	1.08
20	10.00	112.40	0.71
21	10.50	113.40	0.88
22	11.00	114.40	0.87

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(114.4 - 23.97574) (1)/(1) = +9.042E+01 g/L Carbon
 (114.4 - 23.97574) (1)/(1)(12) = +7.535E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3851

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02983 DUP

Date: 12/03/98

Time: 16:51:09

Sample Size = 1.0L

Dil Factor = 1

Blank ID # =

Blank Value = .58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.00	0.00
2	1.01	0.10	100.00
3	1.51	7.30	98.63
4	2.01	43.10	83.06
5	2.51	117.00	63.16
6	3.00	182.60	35.93
7	3.50	223.60	18.34
8	4.00	249.70	10.45
9	4.50	268.30	6.93
10	5.00	281.90	4.82
11	5.50	289.50	2.63
12	6.00	291.90	0.82
13	6.50	293.20	0.44
14	7.00	294.40	0.41
15	7.50	295.60	0.41
16	8.00	296.80	0.40
17	8.50	297.70	0.30
18	9.00	299.00	0.43
19	9.50	299.80	0.27
20	10.00	301.10	0.43
21	10.50	301.90	0.26
22	11.00	303.10	0.40

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(303.1 - 6.380424) (1) / (1) =	+2.967E+02	g/L Carbon
(303.1 - 6.380424) (1) / (1) (12) =	+2.473E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

3852

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02983 DUP

Date: 12/03/98

Time: 17:05:52

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.00	0.00
2	1.00	1.10	100.00
3	1.50	16.10	93.17
4	2.00	42.40	62.03
5	2.50	70.10	39.51
6	3.00	92.50	24.22
7	3.50	109.90	15.83
8	4.00	123.70	11.16
9	4.50	134.40	7.96
10	5.00	138.90	3.24
11	5.50	141.50	1.84
12	6.00	143.50	1.39
13	6.50	145.10	1.10
14	7.00	146.50	0.96
15	7.50	148.00	1.01
16	8.00	149.50	1.00
17	8.50	150.80	0.86
18	9.00	152.00	0.79
19	9.50	153.50	0.98
20	10.00	154.50	0.65
21	10.50	156.10	1.02
22	11.00	157.20	0.70

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(157.2 - 23.97721) (1)/(1) = +1.332E+02 g/L Carbon
 (157.2 - 23.97721) (1)/(1) (12) = +1.110E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3853

SS = 0.0698g

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02983 TRP , Date: 12/03/98

Time: 17:23:37

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .58 ug/minute C

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.00	0.00
2	1.01	0.20	100.00
3	1.50	5.00	96.00
4	2.00	18.40	72.83
5	2.50	46.30	60.26
6	3.00	91.40	49.34
7	3.50	130.50	29.96
8	4.00	158.30	17.56
9	4.50	179.70	11.91
10	5.00	197.10	8.83
11	5.50	212.00	7.03
12	6.00	224.80	5.69
13	6.50	236.20	4.83
14	7.00	244.60	3.43
15	7.50	249.70	2.04
16	8.00	251.70	0.79
17	8.50	252.80	0.44
18	9.00	253.50	0.28
19	9.50	255.20	0.67
20	10.00	256.00	0.31
21	10.50	257.00	0.39
22	11.00	258.30	0.50

USER INPUT BLANK VALUE

BLANK VALUE = 6.379327 micrograms carbon

BLANK FACTOR = 6.379327 / 10.99884 = +5.8E-01 ug/min Carbon

SAMPLE RESULTS:

(258.3 - 6.37823) (1)/(1) = +2.519E+02 g/L Carbon
 (258.3 - 6.37823) (1)/(1) (12) = +2.099E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0634g

3854

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02983 TRP

Date: 12/03/98

Time: 17:37:31

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.18 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.00	0.00
2	1.01	0.40	100.00
3	1.51	9.60	95.83
4	2.00	23.20	58.62
5	2.50	39.10	40.66
6	3.00	55.00	28.91
7	3.50	69.80	21.20
8	4.00	83.30	16.21
9	4.50	95.40	12.68
10	5.00	106.50	10.42
11	5.50	116.00	8.19
12	6.00	122.20	5.07
13	6.50	124.90	2.16
14	7.00	126.70	1.42
15	7.50	127.80	0.86
16	8.00	129.00	0.93
17	8.50	130.60	1.23
18	9.00	131.90	0.99
19	9.50	132.40	0.38
20	10.00	133.60	0.90
21	10.50	134.50	0.67
22	11.00	135.90	1.03

USER INPUT BLANK VALUE

BLANK VALUE = 23.97747 micrograms carbon

BLANK FACTOR = 23.97747 / 10.99884 = +2.2E+00 ug/min Carbon

SAMPLE RESULTS:

(135.9 - 23.97548) (1)/(1) = +1.119E+02 g/L Carbon
 (135.9 - 23.97548) (1)/(1) (12) = +9.327E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3855

LABCORE Completed Worklist Report for Worklist# 27394

Analyst: jds

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY105 FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0		@TICTOC1 TIC-02	LIQUID	1	1.00E-1	0.100	ug/mL
1 BLNK	0		@TICTOC1 TOC-02	LIQUID	1	1.80E+0	1.800	ug/mL
2 STD	0		@TICTOC1 TIC-02	LIQUID	6.02E+02	5.45E+2	90.532	% Recovery
2 STD	0		@TICTOC1 TOC-02	LIQUID	3.00E+03	2.65E+3	88.333	% Recovery
3 SAMPLE	S98T002969	0	@TICTOC1 TIC-02	LIQUID	N/A	3.85E+03	5.000	ug/mL
3 SAMPLE	S98T002969	0	@TICTOC1 TOC-02	LIQUID	N/A	2.98E+03	40.000	ug/mL
4 DUP	S98T002969	0	@TICTOC1 TIC-02	LIQUID	3.85E+3	3.18E+3	19.061	RPD
4 DUP	S98T002969	0	@TICTOC1 TOC-02	LIQUID	2.98E+3	2.96E+3	0.673	RPD

Final page for worklist# 27394

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

12/02/98 15:10
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27394 Analyst: JS Instrument: CARB2 Book# 25N/2D
26A/2FMethod: LA-342-100 Rev/Mod K-1

Worklist Comment: BY105 FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	LIQUID		
2 STD			@TICTOC1	LIQUID		
3 SAMPLE	S98T002969 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002969 0		@TICTOC1	LIQUID		

Final page for worklist # 27394
JS 12/4/98
 Analyst Signature Date

Mary Jones 12-7-98
 Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3857

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	9.9	38.7
27394	µg of Carbon from Baseline (C2)	9.8	36.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98006015			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/04/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:50 PM		5	40
Sample Point	µg of Carbon	1.00E-01	1.80E+00
BY 105			

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3858

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	554.6	567.1
27394	μg of Carbon from Baseline (C2)	9.8	36.9
Test Code	Standard Book Number	25N12D	26N12F
@TIC/TOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000

Matrix
LIQUIDBatch Number
98006015Rerun
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
MFChemist
MJLAnalyst
JDSDate Complete
12/07/98Analysis Date
12/04/98Analysis Time
04:50 PMSample Point
BY 105QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$)QC Found in $\mu\text{g}/\text{mL}$ = (C1 - C2) * DF / SSQC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if C1 < C2QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
04:50 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.45E+02	2.65E+03
BY 105	Percent Standard Recovery	90.5	88.4

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>MF</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3859

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Size in mL (SS)	TIC	TOC
SAMPLE	Dilution Factor (DF)	0.1000	0.1000
WorkList	µg of Carbon in Sample (C1)	1	1
27394	µg of Carbon from Baseline (C2)	395	335
Test Code		9.8	36.9
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98006015			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T002969			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/04/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:50 PM		5	40
Sample Point	µg of Carbon/mL	3.85E+03	2.98E+03
BY 105			

$\mu\text{g of Carbon/mL} = (C1 - C2) * DF / SS$
 $\mu\text{g of Carbon/mL for TIC} = 5 \text{ if } C1 < C2$
 $\mu\text{g of Carbon/mL for TOC} = 40 \text{ if } C1 < C2$

Data Entered By:	MF	Date:	12/07/98
Signature of Chemist:	<i>NA</i>	Date:	

SAMPLE.WB1 REV 1.0

342100ML

3860

I:\342100\OUT\27394.WB1

12/07/98

09:41:52

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	327.4	333.2
27394	µg of Carbon from Baseline (C2)	9.8	36.9
Test Code	Known µg of C from Original Sample	3.85E+3	2.98E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98006015			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002969			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
JDS			
Date Complete			
12/07/98			
Analysis Date			
12/04/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:50 PM			
Sample Point	µg of Carbon/mL	3.18E+03	2.96E+03
BY 105			

Data Entered By:

MF

Date: 12/07/98

Signature of Chemist:

NA

Date:

SAMPLE.WB1 REV 1.0

342100ML

3861

I:\342100\OUT\27394.WB1

12/07/98

09:42:36

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE

Date: 12/04/98

Time: 10:21:30

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : JD SPELLMAN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.70	100.00
3	1.01	1.80	61.11
4	1:50	2.90	37.93
5	2.00	3.90	25.64
6	2.50	4.50	13.33
7	3.00	4.80	6.25
8	3.50	5.30	9.43
9	4.00	5.50	3.64
10	4.50	5.70	3.51
11	5.00	5.90	3.39
12	5.50	6.20	4.84
13	6.00	6.50	4.62
14	6.50	6.80	4.41
15	7.00	7.10	4.23
16	7.50	7.40	4.05
17	8.00	7.80	5.13
18	8.50	8.10	3.70
19	9.00	8.50	4.71
20	9.50	9.20	7.61
21	10.00	9.20	0.00
22	10.50	9.80	6.12

BLANK VALUE = 9.8 micrograms carbon
 BLANK FACTOR = 9.8 / 10.49805 =

+9.3E-01 ug/min Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3862 TO 3871

Sample Run By:

JD SPELLMAN

22222

3862

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE

Date: 12/04/98

Time: 10:41:45

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE
 Blank Value = N/A

Analyst : JD SPELLMAN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.90	100.00
3	1.01	5.40	83.33
4	1.51	12.80	57.81
5	2.00	19.40	34.02
6	2.51	23.60	17.80
7	3.00	26.70	11.61
8	3.50	28.70	6.97
9	4.00	29.40	2.38
10	4.50	30.00	2.00
11	5.00	30.70	2.28
12	5.50	31.50	2.54
13	6.00	32.30	2.48
14	6.50	32.80	1.52
15	7.00	33.40	1.80
16	7.50	33.80	1.18
17	8.00	34.50	2.03
18	8.50	34.90	1.15
19	9.00	35.40	1.41
20	9.50	36.00	1.67
21	10.00	36.50	1.37
22	10.50	36.90	1.08

BLANK VALUE = 36.9 micrograms carbon

BLANK FACTOR = 36.9 / 10.49902 =

+3.51E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3863

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 12/04/98

Time: 10:58:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .93 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	41.80	100.00
3	1.01	64.20	34.89
4	1.51	99.10	35.22
5	2.00	200.50	50.57
6	2.50	320.30	37.40
7	3.00	418.20	23.41
8	3.50	479.30	12.75
9	4.00	513.90	6.73
10	4.50	531.00	3.22
11	5.00	538.40	1.37
12	5.50	542.00	0.66
13	6.00	544.30	0.42
14	6.50	545.80	0.27
15	7.00	547.10	0.24
16	7.50	548.20	0.20
17	8.00	549.30	0.20
18	8.50	550.40	0.20
19	9.00	551.40	0.18
20	9.50	552.50	0.20
21	10.00	553.80	0.23
22	10.50	554.60	0.14

USER INPUT BLANK VALUE

BLANK VALUE = 9.764091 micrograms carbon

BLANK FACTOR = 9.764091 / 10.49902 = +9.3E-01 ug/min Carbon

SAMPLE RESULTS:

(554.6 - 9.763297) (1)/(1) =

+5.448E+02 g/L Carbon

(554.6 - 9.763297) (1)/(1),(12) =

+4.540E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3864

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD

Date: 12/04/98

Time: 14:14:08

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.51 ug/minute C

Analyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.60	100.00
3	1.01	0.80	25.00
4	1.51	1.70	52.94
5	2.00	2.70	37.04
6	2.50	19.80	86.36
7	3.00	115.70	82.89
8	3.50	261.10	55.69
9	4.00	408.30	36.05
10	4.50	490.70	16.79
11	5.00	527.50	6.98
12	5.50	543.40	2.93
13	6.00	550.60	1.31
14	6.50	554.20	0.65
15	7.00	556.50	0.41
16	7.50	558.90	0.43
17	8.00	560.10	0.21
18	8.50	562.30	0.39
19	9.00	563.30	0.18
20	9.50	564.50	0.21
21	10.00	565.80	0.23
22	10.50	567.10	0.23

USER INPUT BLANK VALUE

BLANK VALUE = 36.85157 micrograms carbon

BLANK FACTOR = 36.85157 / 10.49902 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

{ 567.1 - 36.85157 } (1)/(1) = +5.302E+02 g/L Carbon
{ 567.1 - 36.85157 } (1)/(1)(12) = +4.419E+01 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3865

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 12/04/98

Time: 14:42:55

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .93 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.10	100.00
3	1.01	0.30	66.67
4	1.50	1.60	81.25
5	2.00	2.30	30.43
6	2.50	3.40	32.35
7	3.00	4.00	15.00
8	3.50	4.70	14.89
9	4.00	5.40	12.96
10	4.50	6.00	10.00
11	5.00	6.70	10.45
12	5.50	7.40	9.46
13	6.00	8.10	8.64
14	6.50	8.60	5.81
15	7.00	8.60	0.00
16	7.50	8.60	0.00
17	8.00	8.60	0.00
18	8.50	8.70	1.15
19	9.00	9.00	3.33
20	9.50	9.30	3.23
21	10.00	9.60	3.13
22	10.50	9.90	3.03

USER INPUT BLANK VALUE

BLANK VALUE = 9.764091 micrograms carbon

BLANK FACTOR = 9.764091 / 10.49902 = +9.3E-01 ug/min Carbon

SAMPLE RESULTS:

(9.9 - 9.763978) (1)/(1) =

+1.4E-01 g/L Carbon

(9.9 - 9.763978) (1)/(1) (12) =

+1.1E-02 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3866

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK

Date: 12/04/98

Time: 15:09:54

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.51 ug/minute C

Analyst : JD SPELLMAN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	2.40	100.00
3	1.01	6.90	65.22
4	1.51	11.20	38.39
5	2.00	14.50	22.76
6	2.50	16.80	13.69
7	3.00	18.70	10.16
8	3.50	20.20	7.43
9	4.00	21.80	7.34
10	4.50	23.10	5.63
11	5.00	24.50	5.71
12	5.50	25.90	5.41
13	6.00	27.20	4.78
14	6.50	28.50	4.56
15	7.00	29.90	4.68
16	7.50	31.00	3.55
17	8.00	32.40	4.32
18	8.50	33.60	3.57
19	9.00	34.90	3.72
20	9.50	36.10	3.32
21	10.00	37.40	3.48
22	10.50	38.70	3.36

USER INPUT BLANK VALUE

BLANK VALUE = 36.85157 micrograms carbon

BLANK FACTOR = 36.85157 / 10.49902 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(38.7 - 36.85157) (1)/(1) = +1.85E+00 g/L Carbon
 (38.7 - 36.85157) (1)/(1) (12) = +1.54E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3867

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002969

Date: 12/04/98

Time: 16:06:01

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .93 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.50	100.00
3	1.01	1.50	66.67
4	1.51	6.90	78.26
5	2.00	20.80	66.83
6	2.51	51.90	59.92
7	3.00	99.00	47.58
8	3.50	144.40	31.44
9	4.01	180.80	20.13
10	4.50	209.20	13.58
11	5.00	233.20	10.29
12	5.50	254.30	8.30
13	6.00	273.40	6.99
14	6.50	290.80	5.98
15	7.00	307.00	5.28
16	7.50	322.10	4.69
17	8.00	336.10	4.17
18	8.50	349.20	3.75
19	9.00	361.60	3.43
20	9.50	373.30	3.13
21	10.00	384.40	2.89
22	10.50	395.00	2.68

USER INPUT BLANK VALUE

BLANK VALUE = 9.764091 micrograms carbon

BLANK FACTOR = 9.764091 / 10.49902 = +9.3E-01 ug/min Carbon

SAMPLE RESULTS:

(395 - 9.765) (1)/(1)	=	+3.852E+02	g/L Carbon
(395 - 9.765) (1)/(1) (12)	=	+3.210E+01	Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3868

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002969

Date: 12/04/98

Time: 16:18:40

Sample Size = 1 uL

Analyst : JD SPELLMAN

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 3.51 ug/minute C

% Difference = 10

== Reading	==== Analysis	Time	==== Coulometer	==== % Difference ==
1	0.08		0.00	0.00
2	0.50		0.40	100.00
3	1.01		7.50	94.67
4	1.50		19.50	61.54
5	2.00		38.00	48.68
6	2.50		64.30	40.90
7	3.00		90.80	29.19
8	3.50		114.80	20.91
9	4.00		136.60	15.96
10	4.50		156.80	12.88
11	5.00		175.70	10.76
12	5.50		193.50	9.20
13	6.00		210.40	8.03
14	6.50		226.50	7.11
15	7.00		241.90	6.37
16	7.50		256.70	5.77
17	8.00		270.90	5.24
18	8.50		284.60	4.81
19	9.00		297.80	4.43
20	9.50		310.60	4.12
21	10.00		323.00	3.84
22	10.50		335.00	3.58

USER INPUT BLANK VALUE

BLANK VALUE = 36.85157 micrograms carbon

BLANK FACTOR = 36.85157 / 10.49902 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(335 - 36.84857) (1)/(1) = +2.982E+02 g/L Carbon
 (335 - 36.84857) (1)/(1) (12) = +2.485E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3869

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002969DUP

Date: 12/04/98

Time: 16:31:17

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .93 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	10.60	100.00
3	1.00	21.00	49.52
4	1.51	31.40	33.12
5	2.00	42.30	25.77
6	2.50	55.70	24.06
7	3.00	72.20	22.85
8	3.50	91.00	20.66
9	4.04	110.70	17.80
10	4.54	130.30	15.04
11	5.04	149.30	12.73
12	5.54	167.60	10.92
13	6.04	185.30	9.55
14	6.54	202.60	8.54
15	7.03	219.50	7.70
16	7.53	235.80	6.91
17	8.03	251.90	6.39
18	8.53	267.60	5.87
19	9.03	283.00	5.44
20	9.53	298.10	5.07
21	10.03	312.90	4.73
22	10.53	327.40	4.43

USER INPUT BLANK VALUE

BLANK VALUE = 9.764091 micrograms carbon

BLANK FACTOR = 9.764091 / 10.49902 = +9.3E-01 ug/min Carbon

SAMPLE RESULTS:

(327.4 - 9.794744) (1)/(1) = +3.176E+02 g/L Carbon
 (327.4 - 9.794744) (1)/(1) (12) = +2.647E+01 Molar Carbon

Sample Run By:

JD SPELLMAN

22222

3870

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002969DUP

Date: 12/04/98

Time: 16:56:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.51 ug/minute C

Analyst : JD SPELLMAN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.08	0.00	0.00
2	0.51	13.20	100.00
3	1.01	26.80	50.75
4	1.51	41.30	35.11
5	2.00	57.80	28.55
6	2.50	75.40	23.34
7	3.00	93.00	18.92
8	3.50	110.60	15.91
9	4.00	128.10	13.66
10	4.50	145.40	11.90
11	5.00	162.50	10.52
12	5.50	179.30	9.37
13	6.00	196.00	8.52
14	6.50	212.30	7.68
15	7.00	228.50	7.09
16	7.50	244.20	6.43
17	8.00	259.50	5.90
18	8.50	274.60	5.50
19	9.00	289.50	5.15
20	9.50	304.20	4.83
21	10.00	318.80	4.58
22	10.50	333.20	4.32

USER INPUT BLANK VALUE

BLANK VALUE = 36.85157 micrograms carbon

BLANK FACTOR = 36.85157 / 10.49902 = +3.5E+00 ug/min Carbon

SAMPLE RESULTS:

(333.2 - 36.85157) (1)/(1) = +2.963E+02 g/L Carbon
 (333.2 - 36.85157) (1)/(1) (12) = +2.470E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

JD SPELLMAN

22222

3871

LABCORE Completed Worklist Report for Worklist# 27629

Analyst: pjm

Instrument: CARB1

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 @TICTOC1 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0		@TICTOC1 TIC-02	LIQUID	1	1.70E+0	1.700	ug/mL
1 BLNK	0		@TICTOC1 TOC-02	LIQUID	1	9.50E+0	9.500	ug/mL
2 STD	0		@TICTOC1 TIC-02	LIQUID	6.02E+02	5.96E+2	99.003	% Recovery
2 STD	0		@TICTOC1 TOC-02	LIQUID	3.00E+03	2.72E+3	90.667	% Recovery
3 SAMPLE	S98T003385	0	@TICTOC1 TIC-02	LIQUID	N/A	4.52E+03	5.000	ug/mL
3 SAMPLE	S98T003385	0	@TICTOC1 TOC-02	LIQUID	N/A	2.28E+03	40.000	ug/mL
4 DUP	S98T003385	0	@TICTOC1 TIC-02	LIQUID	4.52E+3	4.44E+3	1.786	RPD
4 DUP	S98T003385	0	@TICTOC1 TOC-02	LIQUID	2.28E+3	2.28E+3	0.000	RPD
5 SPK	S98T003385	0	@TICTOC1 TIC-02	LIQUID	1.00E+02	9.94E+01	99.400	% Recovery
5 SPK	S98T003385	0	@TICTOC1 TOC-02	LIQUID	1.00E+02	9.60E+01	96.000	% Recovery

Final page for worklist# 27629

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

12/18/98 08:10
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27629

Analyst: RW Instrument: CARB2 _____ Book# TIC 25N12DMethod: LA-342-100 Rev/Mod F-1TOC 26N12F

Worklist Comment: BY-105 @TICTOC1 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	LIQUID		
2 STD			@TICTOC1	LIQUID		
3 SAMPLE	S98T003385 0		@TICTOC1	LIQUID	98000437	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T003385 0		@TICTOC1	LIQUID		
5 SPK	S98T003385 0		@TICTOC1	LIQUID		

Final page for worklist # 27629

R. McCann 12/19/98
Analyst Signature DateC. J. Davis 12/21/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3873

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	4	24.3
27629	µg of Carbon from Baseline (C2)	5.7	33.8
Test Code			
@TIC/TOC1			
Matrix			
LIQUID			
Batch Number			
98006253			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/21/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:30 AM		5	40
Sample Point	µg of Carbon	1.70E+00	9.50E+00
BLNK			

Data Entered By:	CJO	Date:	12/21/98
Signature of Chemist:	<i>N/A</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3874

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in µg (C1)	601.5	577.4
27629	µg of Carbon from Baseline (C2)	5.7	33.8
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value (µg/ml)	602	3000
Matrix	QC Actual in µg/mL = Standard Value (µg/mL) QC Found in µg/mL = (C1 - C2) * DF / SS QC Found in µg/mL for TIC = 5 if C1 < C2 QC Found in µg/mL for TOC = 40 if C1 < C2 % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98006253			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/21/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:30 AM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.96E+02	2.72E+03
BLNK	Percent Standard Recovery	99.0	90.6

Data Entered By:	CJO	Date:	12/21/98
Signature of Chemist:	<i>A/A</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3875

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Size in mL (SS)	TIC	TOC
SAMPLE	Dilution Factor (DF)	0.1000	0.1000
Work List	µg of Carbon in Sample (C1)	1	1
27629	µg of Carbon from Baseline (C2)	457.3	261.7
Test Code		5.7	33.8
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98006253			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003385			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/21/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
04:30 AM		5	40
Sample Point	µg of Carbon/mL	4.52E+03	2.28E+03
BLNK			

Data Entered By:	CJO	Date:	12/21/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	449.6	261.3
27629	µg of Carbon from Baseline (C2)	5.7	33.8
Test Code	Known µg of C from Original Sample	4.52E+3	2.28E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
98006253			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003385			
Instrument Code			
CARB1			
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/21/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in µg/mL	5	40
04:30 AM			
Sample Point	µg of Carbon/mL	4.44E+03	2.28E+03
BLNK			

Data Entered By:	CJO	Date:	12/21/98
Signature of Chemist:	<i>N/A</i>	Date:	

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Volume in mL (SS)	0.1000	0.1000
Work List	Final Coulometer Reading in µg (C1)	457.3	261.7
27629	Spiked Vial Data		
Test Code	Sample Volume in mL (SPK SS)	0.1000	0.1000
@TICTOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	756.6	549.8
LIQUID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/ml (SPK CONC)	602	3000
98006253	µg C in baseline (BL)	5.7	33.8
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T003385			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB1	QC Found in µg/mL = (Percent Spike Recovery)*(QC Actual) / 100		
Prepared By			
CJO			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
12/21/98			
Analysis Date			
12/19/98			
Analysis Time			
04:30 AM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.99E+02	2.88E+03
BLNK	Percent Spike Recovery	99.4	96.0

Data Entered By:	CJO	Date:	12/21/98
Signature of Chemist:	NA	Date:	

SPIKE.WB1 REV 1.3

342100ML

3878

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-1

Date: 12/19/98

Time: 00:22:05

Sample Size = 1000 uL
 Dil Factor = 1
 Blank ID # =
 Blank Value = .52 ug/minute C

Analyst : PJ MCCOWN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.50	40.00
3	1.50	0.80	37.50
4	2.00	1.10	27.27
5	2.50	2.90	62.07
6	3.00	21.90	86.76
7	3.50	85.40	74.36
8	4.00	186.30	54.16
9	4.50	302.80	38.47
10	5.00	404.10	25.07
11	5.50	480.30	15.87
12	6.00	531.10	9.57
13	6.50	561.20	5.36
14	7.00	578.60	3.01
15	7.50	587.70	1.55
16	8.00	592.70	0.84
17	8.50	595.80	0.52
18	9.00	597.70	0.32
19	9.50	599.00	0.22
20	10.00	600.00	0.17
21	10.50	600.70	0.12
22	11.00	601.50	0.13

USER INPUT BLANK VALUE

BLANK VALUE = -743.081 micrograms carbon

BLANK FACTOR = -743.081 / -1429.002 = +5.2E-01 ug/min Carbon

SAMPLE RESULTS:

(601.5 - 5.718873) (1)/(1000) = +5.958E-01 g/L Carbon
 (601.5 - 5.718873) (1)/(1000) (12) = +4.965E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3879 TO 3890

Sample Run By: PJ MCCOWN 12/19/98

PJ MCCOWN

00000

1-OMC 25N12D

3879

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-1

Date: 12/19/98

Time: 00:45:14

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.40	0.00
2	1.00	0.80	50.00
3	1.50	1.10	27.27
4	2.00	1.50	26.67
5	2.50	2.00	25.00
6	3.00	2.70	25.93
7	3.50	7.30	63.01
8	4.00	29.20	75.00
9	4.50	88.60	67.04
10	5.00	190.30	53.44
11	5.50	304.20	37.44
12	6.00	401.00	24.14
13	6.50	468.50	14.41
14	7.00	511.70	8.44
15	7.50	537.90	4.87
16	8.00	553.10	2.75
17	8.50	562.30	1.64
18	9.00	567.90	0.99
19	9.50	571.60	0.65
20	10.00	574.10	0.44
21	10.50	576.00	0.33
22	11.00	577.40	0.24

USER INPUT BLANK VALUE

BLANK VALUE = -4387.036 micrograms carbon

BLANK FACTOR = -4387.036 / -1429.002 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(577.4 - 33.76334) (1)/(200) = +2.718E+00 g/L Carbon
 (577.4 - 33.76334) (1)/(200) (12) = +2.265E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN 12/19/98
 PJ MCCOWN 00000

.200ML 26N12F

3880

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE-1

Date: 12/18/98

Time: 23:47:58

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE-1

Blank Value = N/A

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	% Difference ==
1	0.51	0.20	0.00
2	1.01	0.60	66.67
3	1.51	0.80	25.00
4	2.00	1.00	20.00
5	2.50	1.40	28.57
6	3.00	1.60	12.50
7	3.50	1.80	11.11
8	4.00	2.10	14.29
9	4.50	2.30	8.70
10	5.00	2.60	11.54
11	5.50	2.80	7.14
12	6.00	3.10	9.68
13	6.50	3.40	8.82
14	7.00	3.70	8.11
15	7.50	3.90	5.13
16	8.00	4.20	7.14
17	8.50	4.40	4.55
18	9.00	4.70	6.38
19	9.50	5.00	6.00
20	10.00	5.20	3.85
21	10.50	5.50	5.45
22	11.00	5.70	3.51

BLANK VALUE = 5.7 micrograms carbon

BLANK FACTOR = 5.7 / 10.99805 =

+5.2E-01

ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

3881

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE-1

Date: 12/19/98

Time: 00:00:38

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE-1
 Blank Value = N/A

Analyst : PJ MCCOWN
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference
1	0.51	0.20	0.00
2	1.00	0.50	60.00
3	1.50	0.80	37.50
4	2.00	1.10	27.27
5	2.50	2.30	52.17
6	3.00	5.40	57.41
7	3.50	10.60	49.06
8	4.00	15.90	33.33
9	4.50	19.60	18.88
10	5.00	22.40	12.50
11	5.50	24.40	8.20
12	6.00	25.60	4.69
13	6.50	26.40	3.03
14	7.00	27.20	2.94
15	7.50	28.20	3.55
16	8.00	29.40	4.08
17	8.50	30.50	3.61
18	9.00	31.50	3.17
19	9.50	32.20	2.17
20	10.00	32.90	2.13
21	10.50	33.40	1.50
22	%-1429.00	33.80	1.18

BLANK VALUE = 33.8 micrograms carbon
 BLANK FACTOR = 33.8 / -1429.002 =

-2.37E-02 ug/min Carbon

11-0
 RM
 12/18/98

3.07

Sample Run By:

PJ MCCOWN

00000

3882

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-1

Date: 12/19/98

Time: 02:47:21

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = .52 ug/minute C

Analyst ; PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.20	0.00
2	1.01	0.40	50.00
3	1.51	0.50	20.00
4	2.00	0.70	28.57
5	2.50	0.90	22.22
6	3.00	1.10	18.18
7	3.50	1.30	15.38
8	4.00	1.50	13.33
9	4.50	1.60	6.25
10	5.00	1.80	11.11
11	5.50	1.90	5.26
12	6.00	2.10	9.52
13	6.50	2.30	8.70
14	7.00	2.50	8.00
15	7.50	2.60	3.85
16	8.00	2.90	10.34
17	8.50	3.20	9.37
18	9.00	3.30	3.03
19	9.50	3.50	5.71
20	10.00	3.70	5.41
21	10.50	3.90	5.13
22	11.00	4.00	2.50

USER INPUT BLANK VALUE

BLANK VALUE = -743.081 micrograms carbon

BLANK FACTOR = -743.081 / -1429.002 = +5.2E-01 ug/min Carbon

SAMPLE RESULTS:

(4 - 5.719389) (1) / (1) =
(4 - 5.719389) (1) / (1) (12) =

< 5.00 E-3 g/L Carbon
< 4.17 E-4 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

3883

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-1

Date: 12/19/98

Time: 03:01:04

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.20	0.00
2	1.01	0.50	60.00
3	1.50	0.90	44.44
4	2.00	2.10	57.14
5	2.50	4.90	57.14
6	3.00	9.70	49.48
7	3.50	13.90	30.22
8	4.00	16.70	16.77
9	4.50	18.60	10.22
10	5.00	19.70	5.58
11	5.50	20.40	3.43
12	6.00	21.00	2.86
13	6.50	21.50	2.33
14	7.00	21.90	1.83
15	7.50	22.30	1.79
16	8.00	22.60	1.33
17	8.50	22.90	1.31
18	9.00	23.20	1.29
19	9.50	23.50	1.28
20	10.00	23.80	1.26
21	10.50	24.00	0.83
22	11.00	24.30	1.23

USER INPUT BLANK VALUE

BLANK VALUE = -4387.036 micrograms carbon

BLANK FACTOR = -4387.036 / -1429.002 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(24.3 - 33.76082) (1)/(1) = < 5.00 E-3 g/L Carbon
 (24.3 - 33.76082) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3884

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003385

Date: 12/19/98

Time: 03:13:54

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .52 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.20	0.00
2	1.04	0.50	60.00
3	1.54	0.70	28.57
4	2.04	2.10	66.67
5	2.54	14.50	85.52
6	3.04	62.50	76.80
7	3.54	145.90	57.16
8	4.04	251.40	41.96
9	4.54	342.20	26.53
10	5.04	398.40	14.11
11	5.54	426.60	6.61
12	6.04	439.60	2.96
13	6.54	445.70	1.37
14	7.04	449.00	0.73
15	7.53	451.00	0.44
16	8.03	452.50	0.33
17	8.53	453.50	0.22
18	9.03	454.40	0.20
19	9.53	455.20	0.18
20	10.03	456.00	0.18
21	10.53	456.70	0.15
22	11.03	457.30	0.13

USER INPUT BLANK VALUE

BLANK VALUE = -743.081 micrograms carbon

BLANK FACTOR = -743.081 / -1429.002 = +5.2E-01 ug/min Carbon

SAMPLE RESULTS:

(457.3 - 5.737075) (1)/(1)	=	+4.516E+02	g/L Carbon
(457.3 - 5.737075) (1)/(1) (12)	=	+3.763E+01	Molar Carbon

Sample Run By: *PJ*

PJ MCCOWN

00000

100 mL

3885

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003385

Date: 12/19/98

Time: 03:28:03

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.20	41.67
3	1.50	2.00	40.00
4	2.00	6.70	70.15
5	2.50	28.30	76.33
6	3.00	81.60	65.32
7	3.50	146.20	44.19
8	4.00	193.00	24.25
9	4.50	219.90	12.23
10	5.00	234.30	6.15
11	5.50	242.50	3.38
12	6.00	247.80	2.14
13	6.50	251.10	1.31
14	7.00	253.40	0.91
15	7.50	254.90	0.59
16	8.00	256.20	0.51
17	8.50	257.40	0.47
18	9.00	258.50	0.43
19	9.50	259.50	0.39
20	10.00	260.40	0.35
21	10.50	261.10	0.27
22	11.00	261.70	0.23

USER INPUT BLANK VALUE

BLANK VALUE = -4387.036 micrograms carbon

BLANK FACTOR = -4387.036 / -1429.002 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(261.7 - 33.76077) (1)/(1) = +2.279E+02 g/L Carbon
 (261.7 - 33.76077) (1)/(1) (12) = +1.899E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

3886

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3385 DUP

Date: 12/19/98

Time: 03:40:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .52 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time. ==== Coulometer ==== % Difference ==

1	0.50	0.40	0.00
2	1.01	0.90	55.56
3	1.50	1.30	30.77
4	2.00	2.50	48.00
5	2.50	14.10	82.27
6	3.00	67.70	79.17
7	3.50	181.50	62.70
8	4.00	298.10	39.11
9	4.50	371.60	19.78
10	5.00	409.40	9.23
11	5.50	426.70	4.05
12	6.00	434.40	1.77
13	6.50	438.50	0.94
14	7.00	440.90	0.54
15	7.50	442.60	0.38
16	8.00	443.90	0.29
17	8.50	445.10	0.27
18	9.00	446.10	0.22
19	9.50	447.00	0.20
20	10.00	448.00	0.22
21	10.50	448.70	0.16
22	11.00	449.60	0.20

USER INPUT BLANK VALUE

BLANK VALUE = -743.081 micrograms carbon

BLANK FACTOR = -743.081 / -1429.002 = +5.2E-01 ug/min Carbon

SAMPLE RESULTS:

(449.6 - 5.71896) (1)/(1) = +4.439E+02 g/L Carbon
 (449.6 - 5.71896) (1)/(1) (12) = +3.699E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.100mL

3887

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3385 DUP

Date: 12/19/98

Time: 03:55:13

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 3.07 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.50	53.33
3	1.50	9.10	83.52
4	2.00	52.80	82.77
5	2.50	133.00	60.30
6	3.00	191.10	30.40
7	3.50	220.50	13.33
8	4.00	234.90	6.13
9	4.50	242.60	3.17
10	5.00	247.40	1.94
11	5.50	250.60	1.28
12	6.00	252.90	0.91
13	6.50	254.60	0.67
14	7.00	256.00	0.55
15	7.50	257.10	0.43
16	8.00	258.00	0.35
17	8.50	258.70	0.27
18	9.00	259.30	0.23
19	9.50	259.80	0.19
20	10.00	260.40	0.23
21	10.50	260.90	0.19
22	11.00	261.30	0.15

USER INPUT BLANK VALUE

BLANK VALUE = -4387.036 micrograms carbon

BLANK FACTOR = -4387.036 / -1429.002 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(261.3 - 33.76082) (1) / (1) = +2.275E+02 g/L Carbon
 (261.3 - 33.76082) (1) / (1) (12) = +1.896E+01 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!!!>>>

Sample Run By: *PJ*

PJ MCCOWN

00000

100mc

3888

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3385 SPK

Date: 12/19/98

Time: 04:09:10

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .52 ug/minute C

Analyst :

PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.60	0.00
2	1.01	1.20	50.00
3	1.50	6.60	81.82
4	2.00	91.30	92.77
5	2.50	297.80	69.34
6	3.00	508.40	41.42
7	3.50	637.40	20.24
8	4.00	699.90	8.93
9	4.50	727.60	3.81
10	5.00	738.50	1.48
11	5.50	743.20	0.63
12	6.00	745.90	0.36
13	6.50	747.70	0.24
14	7.00	749.20	0.20
15	7.50	750.50	0.17
16	8.00	751.50	0.13
17	8.50	752.60	0.15
18	9.00	753.60	0.13
19	9.50	754.30	0.09
20	10.00	755.10	0.11
21	10.50	755.90	0.11
22	11.00	756.60	0.09

USER INPUT BLANK VALUE

BLANK VALUE = -743.081 micrograms carbon

BLANK FACTOR = -743.081 / -1429.002 =

+5.2E-01

ug/min Carbon

SAMPLE RESULTS:

(756.6 - 5.718445) (1)/(1) =

+7.509E+02 g/L Carbon

(756.6 - 5.718445) (1)/(1) (12) =

+6.257E+01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

.106 uL

+ .500 uL 26 N12D

3889

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 3385 SPK

Date: 12/19/98

Time: 04:22:28

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 3.07 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.90	0.00
2	1.01	1.40	35.71
3	1.50	3.00	53.33
4	2.00	17.00	82.35
5	2.50	112.20	84.85
6	3.00	299.80	62.58
7	3.50	429.10	30.13
8	4.00	489.50	12.34
9	4.50	517.20	5.36
10	5.00	529.00	2.23
11	5.50	534.60	1.05
12	6.00	538.10	0.65
13	6.50	540.50	0.44
14	7.00	542.40	0.35
15	7.50	543.80	0.26
16	8.00	545.20	0.26
17	8.50	546.10	0.16
18	9.00	547.00	0.16
19	9.50	547.80	0.15
20	10.00	548.50	0.13
21	10.50	549.10	0.11
22	11.00	549.80	0.13

USER INPUT BLANK VALUE

BLANK VALUE = -4387.036 micrograms carbon

BLANK FACTOR = -4387.036 / -1429.002 = +3.1E+00 ug/min Carbon

SAMPLE RESULTS:

(549.8 - 33.76077) (1)/(1) = +5.160E+02 g/L Carbon
 (549.8 - 33.76077) (1)/(1) (12) = +4.300E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: *PJ*

PJ MCCOWN

00000

-100mL

+ 100mL 26WIDK

3893

LABCORE Completed Worklist Report for Worklist# 27630

Analyst: slh Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY-105 @TICTOC1 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0	0	ETICTOCI TIC-02	SOLID	1	4.50E+0	4.500	ug/g
1 BLNK	0	0	ETICTOCI TOC-02	SOLID	1	1.51E+1	15.100	ug/g
2 STD	0	0	ETICTOCI TIC-02	SOLID	5.02E+02	5.44E+2	90.365	% Recovery
2 STD	0	0	ETICTOCI TOC-02	SOLID	3.00E+03	2.80E+3	93.333	% Recovery
3 SAMPLE	S98T003386	0	ETICTOCI TIC-02	SOLID	N/A	3.19E+03	5.000	ug/g
3 SAMPLE	S98T003386	0	ETICTOCI TOC-02	SOLID	N/A	1.18E+03	40.000	ug/g
4 DUP	S98T003386	0	ETICTOCI TIC-02	SOLID	3.19E+3	3.40E+3	6.373	RPD
4 DUP	S98T003386	0	ETICTOCI TOC-02	SOLID	1.18E+3	1.26E+3	6.557	RPD
5 SPK	S98T003386	0	ETICTOCI TIC-02	SOLID	1.00E+02	1.05E+02	105.000	% Recovery
5 SPK	S98T003386	0	ETICTOCI TOC-02	SOLID	1.00E+02	9.52E+01	95.200	% Recovery
6 SAMPLE	S98T003391	0	ETICTOCI TIC-02	SOLID	N/A	6.17E+03	5.000	ug/g
6 SAMPLE	S98T003391	0	ETICTOCI TOC-02	SOLID	N/A	6.34E+02	40.000	ug/g
7 DUP	S98T003391	0	ETICTOCI TIC-02	SOLID	6.17E+3	5.20E+3	17.062	RPD
7 DUP	S98T003391	0	ETICTOCI TOC-02	SOLID	6.34E+2	6.56E+2	3.411	RPD
8 SPK	S98T003391	0	ETICTOCI TIC-02	SOLID	1.00E+02	8.88E+01	88.800	% Recovery
8 SPK	S98T003391	0	ETICTOCI TOC-02	SOLID	1.00E+02	9.43E+01	94.300	% Recovery
9 SAMPLE	S98T003396	0	ETICTOCI TIC-02	SOLID	N/A	4.30E+03	5.000	ug/g
9 SAMPLE	S98T003396	0	ETICTOCI TOC-02	SOLID	N/A	2.19E+04	40.000	ug/g
10 DUP	S98T003396	0	ETICTOCI TIC-02	SOLID	4.30E+3	5.10E+3	17.021	RPD
10 DUP	S98T003396	0	ETICTOCI TOC-02	SOLID	2.19E+4	2.09E+4	4.673	RPD
11 SPK	S98T003396	0	ETICTOCI TIC-02	SOLID	1.00E+02	1.03E+02	103.000	% Recovery
11 SPK	S98T003396	0	ETICTOCI TOC-02	SOLID	1.00E+02	7.46E+01	74.600	% Recovery

Final page for worklist# 27630

Analyst Signature _____ Date _____

Gregory J. Gray 12-28-98
Analyst Signature _____ Date _____

[Signature] 12/21/98
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27630

Analyst: SLH Instrument: CARB2 Book# TIC 25N12-D
 Method: LA-342-100 Rev/Mod F-1 TOC 26N12-F
 Worklist Comment: BY-105 @TICTOC1 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	SOLID		
2 STD			@TICTOC1	SOLID		
3 SAMPLE	S98T003386 0		@TICTOC1	SOLID	98000490	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02						
4 DUP	S98T003386 0		@TICTOC1	SOLID		
5 SPK	S98T003386 0		@TICTOC1	SOLID		
6 SAMPLE	S98T003391 0		@TICTOC1	SOLID	98000436	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02						
7 DUP	S98T003391 0		@TICTOC1	SOLID		
8 SPK	S98T003391 0		@TICTOC1	SOLID		
9 SAMPLE	S98T003396 0		@TICTOC1	SOLID	98000436	BY-105 (2)
Analytes Requested: TIC-02 , TOC-02						
10 DUP	S98T003396 0		@TICTOC1	SOLID		
11 SPK	S98T003396 0		@TICTOC1	SOLID		

Final page for worklist # 27630

Sandra H. Bond
 Analyst Signature Date
 12-19-98

Mary Franz
 Analyst Signature Date
 12-28-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	ug of Carbon in Sample (C1)	11.6	32.5
27630	ug of Carbon from Baseline (C2)	7.1	17.4

Test Code

@TIC/TOC1

Matrix

SOLID

Batch Number

98006254

Rerun

0

 $\mu\text{g of Carbon} = |C1 - C2|$

Sample Prep

N/A

Sample #

BLNK

Instrument Code

CARB2

Prepared By

MF

Chemist

MJL

Analyst

SLH

Date Complete

12/28/98

Analysis Date

12/19/98

Analysis Time

05:00 PM

Method Detection Limit in ug/g

Sample Point

BY-105

 $\mu\text{g of Carbon}$

TIC

TOC

5

40

4.50E+00

1.51E+01

Data Entered By:

MF

Date: 12/28/98

Signature of Chemist:

Date:

BLANK.WB1 REV 1.0

342100ML

3893

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	551	578.3
27630	μg of Carbon from Baseline (C2)	7.1	17.4
Test Code	Standard Book Number	25N12D	26N12F
@TICTOC1	Standard Value ($\mu\text{g}/\text{ml}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
98006254			
ReRun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
05:00 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	5.44E+02	2.80E+03
BY-105	Percent Standard Recovery	90.3	93.5

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	<i>MF</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3894

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0820	0.0820
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	268.7	114.4
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003386			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 PM			
Sample Point	µg of Carbon/g	3.19E+03	1.18E+03
BY-105			

Data Entered By:

MF

Date: 12/28/98

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3895

I:\342100\OUT\27630.WB1

12/28/98

13:20:56

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0814	0.0814
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	283.5	119.8
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code	Known µg of C from Original Sample	3.19E+3	1.18E+3
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003386			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:00 PM		5	40
Sample Point	µg of Carbon/g	3.40E+03	1.26E+03
BY-105			

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SPIKE5

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0820	0.0820
Work List	Final Coulometer Reading in µg (C1)	268.7	114.4
27630	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0838	0.0838
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	591.7	402
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006254	µg C in baseline (BL)	7.1	17.4
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T003386			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time			
05:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.35E+02	2.85E+03
BY:105	Percent Spike Recovery	105.4	95.2

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	NA	Date:	

SPIKE WB1 REV 1.3

342100ML

3897

WORKBOOK PAGE: SAM6

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0996	0.0996
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	621.9	80.5
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003391			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 PM			
Sample Point	µg of Carbon/g	6.17E+03	6.34E+02
BY-105			

Data Entered By:

MF

Date: 12/28/98

Signature of Chemist:

Date:

WORKBOOK PAGE: DUP7

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0860	0.0860
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	454.3	73.8
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code	Known µg of C from Original Sample	6.17E+3	6.34E+2
@TIC/TOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003391			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	TIC	TOC
05:00 PM		5	40
Sample Point	µg of Carbon/g	5.20E+03	6.56E+02
BY-105			

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: SPIKE8

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0996	0.0996
Work List	Final Coulometer Reading in µg (C1)	621.9	80.5
27630	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0897	0.0897
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in µg (C2)	828.1	357.1
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in µg/mL (SPK CONC)	602	3000
98006254	µg C in baseline (BL)	7.1	17.4
Rerun			
0			
Sample Prep			
N/A			
Sample #	Percent Spike Recovery = ((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100		
S98T003391			
Instrument Code	QC Actual in µg/mL = Spike Value (µg/mL)		
CARB2	QC Found in µg/mL = (Percent Spike Recovery) * (QC Actual) / 100		
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time			
05:00 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	5.35E+02	2.83E+03
BY-105	Percent Spike Recovery	88.8	94.3

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	NA	Date:	

SPIKE.WB1 REV 1.3

342100ML

3900

WORKBOOK PAGE: SAM9

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0339	0.0339
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	152.8	759.4
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code			
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003396			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 PM			
Sample Point	µg of Carbon/g	4.30E+03	2.19E+04
BY-105			

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	<i>NA</i>	Date:	

WORKBOOK PAGE: DUP10

TIC/TOC : LA-342-100 (F-1)

SOLIDS

		TIC	TOC
Type	Sample Size in g (SS)	0.0304	0.0304
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	162.1	651.3
27630	µg of Carbon from Baseline (C2)	7.1	17.4
Test Code	Known µg of C from Original Sample	4.30E+3	2.19E+4
@TICTOC1			
Matrix			
SOLID			
Batch Number			
98006254			
Rerun	µg of Carbon/g = (C1-C2) * DF / SS		
0	µg of Carbon/g for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/g for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T003396			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analysis Date			
12/19/98			
Analysis Time	Method Detection Limit in ug/g	5	40
05:00 PM			
Sample Point	µg of Carbon/g	5.10E+03	2.09E+04
BY-105			

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:		Date:	

WORKBOOK PAGE: SPIKE11

TIC/TOC : LA-342-100 (F-1)

SOLIDS

Type	Sample Vial Data	TIC	TOC
SPK	Sample Size in g (SS)	0.0339	0.0339
Work List	Final Coulometer Reading in μg (C1)	152.8	759.4
27630	Spiked Vial Data		
Test Code	Sample Size in g (SPK SS)	0.0284	0.0284
@TIC/TOC1	Amount of Spike Std. in mL (SPK VOL)	0.500	0.100
Matrix	Final Coulometer Reading in μg (C2)	438.5	862.9
SOLID	Spike Book Number	25N12D	26N12F
Batch Number	Spike Standard Value in $\mu\text{g/mL}$ (SPK CONC)	602	3000
98006254	$\mu\text{g C}$ in baseline (BL)	7.1	17.4
Rerun			
0			
Sample Prep			
N/A			
Sample #			
S98T003396			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
12/28/98			
Analyst Date			
12/19/98			
Analysis Time			
05:00 PM			
Sample Point			
BY-105			

Percent Spike Recovery = $((C2-BL) - (C1-BL) * (SPK SS) / SS) / ((SPK CONC) * (SPK VOL)) * 100$

QC Actual in $\mu\text{g/mL}$ = Spike Value ($\mu\text{g/mL}$)

QC Found in $\mu\text{g/mL}$ = (Percent Spike Recovery) * (QC Actual) / 100

	TIC	TOC
QC Actual in $\mu\text{g/mL}$	6.02E+02	3.00E+03
QC Found in $\mu\text{g/mL}$	6.19E+02	2.24E+03
Percent Spike Recovery	102.8	74.6

Data Entered By:	MF	Date:	12/28/98
Signature of Chemist:	<i>N/A</i>	Date:	

SPIKE.WB1 REV 1.3

342100ML

3903

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/19/98

Time: 09:44:32

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst.: SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.00	1.50	60.00
3	1.50	34.60	95.66
4	2.00	170.00	79.65
5	2.50	329.70	48.44
6	3.00	441.40	25.31
7	3.50	500.90	11.88
8	4.00	527.80	5.10
9	4.50	539.00	2.08
10	5.00	544.40	0.99
11	5.50	547.20	0.51
12	6.00	549.10	0.35
13	6.50	550.50	0.25
14	7.00	551.00	0.09
15	7.50	551.00	0.00
16	8.00	551.00	0.00
17	8.50	551.00	0.00
18	9.00	551.00	0.00
19	9.50	551.00	0.00
20	10.00	551.00	0.00
21	10.50	551.00	0.00
22	11.00	551.00	0.00

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(551 - 7.148691) (1) / (1000) = +5.439E-01 g/L Carbon
 (551 - 7.148691) (1) / (1000) (12) = +4.532E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION ANALYSIS ON PAGES 3904 TO 3927.

Sample Run By:

SL HOOD

00002

1ml 55 25N(2-D)

3904

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 12/19/98

Time: 09:57:12

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	0.90	55.56
3	1.51	9.20	90.22
4	2.00	62.60	85.30
5	2.50	188.70	66.83
6	3.01	352.20	46.42
7	3.51	468.90	24.89
8	4.00	526.50	10.94
9	4.50	550.90	4.43
10	5.00	561.00	1.80
11	5.50	565.60	0.81
12	6.00	568.30	0.48
13	6.50	570.00	0.30
14	7.00	571.50	0.26
15	7.50	572.80	0.23
16	8.00	573.70	0.16
17	8.50	574.50	0.14
18	9.00	575.40	0.16
19	9.50	576.30	0.16
20	10.00	576.90	0.10
21	10.50	577.60	0.12
22	11.00	578.30	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(578.3 - 17.37981) (1) / (200) = +2.805E+00 g/L Carbon
 (578.3 - 17.37981) (1) / (200) (12) = +2.337E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

200ul 26N12-F

3905

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/19/98

Time: 08:46:19

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.20	0.00
2	1.01	0.50	60.00
3	1.51	0.90	44.44
4	2.01	1.40	35.71
5	2.51	1.80	22.22
6	3.01	2.20	18.18
7	3.51	2.50	12.00
8	4.00	2.90	13.79
9	4.50	3.30	12.12
10	5.00	3.50	5.71
11	5.50	3.90	10.26
12	6.00	4.10	4.88
13	6.50	4.40	6.82
14	7.00	4.70	6.38
15	7.50	5.00	6.00
16	8.00	5.30	5.66
17	8.50	5.60	5.36
18	9.00	5.90	5.08
19	9.50	6.20	4.84
20	10.00	6.50	4.62
21	10.50	6.90	5.80
22	11.00	7.10	2.82

BLANK VALUE = 7.1 micrograms carbon

BLANK FACTOR = 7.1 / 10.99884 =

+6.5E-01 ug/min Carbon

Sample Run By:

SL HOOD

00002

3906

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
 TICTOC REV 2.0
 <<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 12/19/98

Time: 09:01:49

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.20	0.00
2	1.00	0.50	60.00
3	1.50	1.70	70.59
4	2.00	4.70	63.83
5	2.50	7.50	37.33
6	3.00	9.40	20.21
7	3.50	10.70	12.15
8	4.00	11.60	7.76
9	4.50	12.20	4.92
10	5.00	12.80	4.69
11	5.50	13.30	3.76
12	6.00	13.80	3.62
13	6.50	14.10	2.13
14	7.00	14.50	2.76
15	7.50	15.00	3.33
16	8.00	15.40	2.60
17	8.50	15.70	1.91
18	9.00	16.00	1.88
19	9.50	16.50	3.03
20	10.00	16.70	1.20
21	10.50	17.00	1.76
22	11.00	17.40	2.30

BLANK VALUE = 17.4 micrograms carbon

BLANK FACTOR = 17.4 / 10.99799 = +1.58E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3907

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/19/98

Time: 10:11:38

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.10	54.55
3	1.51	2.50	56.00
4	2.01	3.70	32.43
5	2.50	4.40	15.91
6	3.00	5.10	13.73
7	3.50	5.70	10.53
8	4.00	6.20	8.06
9	4.50	6.60	6.06
10	5.00	7.00	5.71
11	5.50	7.40	5.41
12	6.00	7.80	5.13
13	6.50	8.20	4.88
14	7.00	8.60	4.65
15	7.50	9.00	4.44
16	8.00	9.30	3.23
17	8.50	9.80	5.10
18	9.00	10.10	2.97
19	9.50	10.50	3.81
20	10.00	10.80	2.78
21	10.50	11.20	3.57
22	11.00	11.60	3.45

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(11.6 - 7.149246) (1)/(1)	=	+4.45E+00	g/L Carbon
(11.6 - 7.149246) (1)/(1)(12)	=	+3.71E-01	Molar Carbon

Sample Run By:

SL HOOD

00002

3908

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 12/19/98

Time: 10:23:42

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 1.58 ug/minute C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.70	57.14
3	1.50	1.40	50.00
4	2.00	4.90	71.43
5	2.50	11.00	55.45
6	3.00	17.40	36.78
7	3.50	21.70	19.82
8	4.00	24.30	10.70
9	4.50	25.70	5.45
10	5.00	26.60	3.38
11	5.50	27.30	2.56
12	6.00	28.00	2.50
13	6.50	28.70	2.44
14	7.00	29.10	1.37
15	7.50	29.70	2.02
16	8.00	30.10	1.33
17	8.50	30.50	1.31
18	9.00	30.90	1.29
19	9.50	31.30	1.28
20	10.00	31.80	1.57
21	10.50	32.10	0.93
22	11.00	32.50	1.23

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(32.5 - 17.37807) (1)/(1) = +1.51E+01 g/L Carbon
 (32.5 - 17.37807) (1)/(1) (12) = +1.26E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3909

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003386

Date: 12/19/98

Time: 10:42:50

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.20	58.33
3	1.50	14.40	91.67
4	2.00	66.80	78.44
5	2.50	133.90	50.11
6	3.00	188.20	28.85
7	3.50	222.30	15.34
8	4.00	241.00	7.76
9	4.50	250.50	3.79
10	5.00	255.20	1.84
11	5.50	258.00	1.09
12	6.00	259.60	0.62
13	6.50	261.10	0.57
14	7.00	262.10	0.38
15	7.50	263.20	0.42
16	8.00	264.00	0.30
17	8.50	264.90	0.34
18	9.00	265.70	0.30
19	9.50	266.50	0.30
20	10.00	267.30	0.30
21	10.50	267.90	0.22
22	11.00	268.70	0.30

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(268.7 - 7.149206) (1)/(1) = +2.616E+02 g/L Carbon
 (268.7 - 7.149206) (1)/(1) (12) = +2.180E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0820g

3910

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S987T003386

Date: 12/19/98

Time: 10:55:21

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.50	0.00
2	1.01	1.00	50.00
3	1.50	1.90	47.37
4	2.00	8.20	76.83
5	2.50	28.30	71.02
6	3.00	54.40	47.98
7	3.50	75.80	28.23
8	4.00	89.30	15.12
9	4.50	96.80	7.75
10	5.00	100.90	4.06
11	5.50	103.50	2.51
12	6.00	105.40	1.80
13	6.50	106.90	1.40
14	7.00	108.20	1.20
15	7.50	109.20	0.92
16	8.00	110.10	0.82
17	8.50	111.00	0.81
18	9.00	111.80	0.72
19	9.50	112.50	0.62
20	10.00	113.20	0.62
21	10.50	113.80	0.53
22	11.00	114.40	0.52

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(114.4 - 17.37817) (1)/(1) = +9.702E+01 g/L Carbon
 (114.4 - 17.37817) (1)/(1) (12) = +8.085E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0820g

3911

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03386 DUP   Date: 12/19/98   Time: 11:07:18

Sample Size = 1 uL   Analyst : SL HOOD
Dil Factor = 1   Min Readings = 22
Blank ID # =   Max Readings = 22
Blank Value = .65 ug/minute C   % Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.60	0.00
2	1.00	1.20	50.00
3	1.50	3.10	61.29
4	2.00	32.90	90.58
5	2.50	97.90	66.39
6	3.00	162.30	39.68
7	3.50	210.20	22.79
8	4.00	240.30	12.53
9	4.50	256.90	6.46
10	5.00	265.80	3.35
11	5.50	270.40	1.70
12	6.00	273.10	0.99
13	6.50	274.90	0.65
14	7.00	276.40	0.54
15	7.50	277.50	0.40
16	8.00	278.50	0.36
17	8.50	279.40	0.32
18	9.00	280.50	0.39
19	9.50	281.10	0.21
20	10.00	281.90	0.28
21	10.50	282.70	0.28
22	11.00	283.50	0.28

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(283.5 - 7.147501) (1) / (1) = +2.764E+02 g/L Carbon
 (283.5 - 7.147501) (1) / (1) (12) = +2.303E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0814g

3912

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03386 DUP Date: 12/19/98 Time: 11:20:23

Sample Size = 1 uL Analyst : SL HOOD
Dil Factor = 1 Min Readings = 22
Blank ID # = Max Readings = 22
Blank Value = 1.58 ug/minute C % Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.90	0.00
2	1.01	1.60	43.75
3	1.51	9.10	82.42
4	2.00	37.20	75.54
5	2.50	66.30	43.89
6	3.00	85.90	22.82
7	3.50	96.70	11.17
8	4.00	102.60	5.75
9	4.50	106.10	3.30
10	5.00	108.10	1.85
11	5.50	110.00	1.73
12	6.00	111.40	1.26
13	6.50	112.60	1.07
14	7.00	113.70	0.97
15	7.50	114.70	0.87
16	8.00	115.50	0.69
17	8.50	116.40	0.77
18	9.00	117.10	0.60
19	9.50	117.80	0.59
20	10.00	118.50	0.59
21	10.50	119.20	0.59
22	11.00	119.80	0.50

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(119.8 - 17.37527) (1)/(1) = +1.024E+02 g/L Carbon
(119.8 - 17.37527) (1)/(1) (12) = +8.535E+00 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0814g

3913

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03386 SPK

Date: 12/19/98

Time: 11:32:23

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.30	46.15
3	1.50	13.40	90.30
4	2.00	116.10	88.46
5	2.50	265.80	56.32
6	3.00	392.40	32.26
7	3.50	478.80	18.05
8	4.00	529.50	9.58
9	4.50	557.20	4.97
10	5.00	570.10	2.26
11	5.50	576.30	1.08
12	6.00	579.50	0.55
13	6.50	581.50	0.34
14	7.00	583.30	0.31
15	7.50	584.60	0.22
16	8.00	585.80	0.20
17	8.50	587.00	0.20
18	9.00	588.00	0.17
19	9.50	589.00	0.17
20	10.00	589.90	0.15
21	10.50	590.80	0.15
22	11.00	591.70	0.15

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(591.7 - 7.149286) (1)/(1)	=	+5.846E+02	g/L Carbon
(591.7 - 7.149286) (1)/(1) (12)	=	+4.871E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

55 = .0838g + .500ml spk 25/12-1

3914

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03386 SPK Date: 12/19/98 Time: 13:35:09

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = 1.58 ug/minute.C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.80	0.00
2	1.01	1.70	52.94
3	1.51	17.50	90.29
4	2.00	87.00	79.89
5	2.50	198.70	56.22
6	3.00	292.70	32.11
7	3.50	344.20	14.96
8	4.00	368.30	6.54
9	4.50	379.10	2.85
10	5.00	384.70	1.46
11	5.50	387.90	0.82
12	6.00	390.30	0.61
13	6.50	392.20	0.48
14	7.00	393.80	0.41
15	7.50	395.20	0.35
16	8.00	396.30	0.28
17	8.50	397.40	0.28
18	9.00	398.50	0.28
19	9.50	399.40	0.23
20	10.00	400.40	0.25
21	10.50	401.20	0.20
22	11.00	402.00	0.20

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(402 - 17.37527) (1)/(1) = +3.846E+02 g/L Carbon
(402 - 17.37527) (1)/(1) (12) = +3.205E+01 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0838g + .100ml spk 26N12-F

3915

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003391

Date: 12/19/98

Time: 14:23:05

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.10	45.45
3	1.50	33.40	96.71
4	2.00	200.80	83.37
5	2.50	350.90	42.78
6	3.00	453.50	22.62
7	3.50	522.00	13.12
8	4.00	567.10	7.95
9	4.50	592.00	4.21
10	5.00	602.90	1.81
11	5.50	608.00	0.84
12	6.00	610.70	0.44
13	6.50	612.70	0.33
14	7.00	614.10	0.23
15	7.50	615.40	0.21
16	8.00	616.50	0.18
17	8.50	617.50	0.16
18	9.00	618.60	0.18
19	9.50	619.40	0.13
20	10.00	620.20	0.13
21	10.50	621.10	0.14
22	11.00	621.90	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(621.9 - 7.147421) (1)/(1)	=	+6.148E+02	g/L Carbon
(621.9 - 7.147421) (1)/(1) (12)	=	+5.123E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0996g

3916

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003391

Date: 12/19/98

Time: 14:35:24

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.70	0.00
2	1.01	1.20	41.67
3	1.51	7.90	84.81
4	2.01	24.40	67.62
5	2.51	40.00	39.00
6	3.00	50.70	21.10
7	3.50	57.40	11.67
8	4.00	61.70	6.97
9	4.50	64.80	4.78
10	5.00	67.20	3.57
11	5.50	69.10	2.75
12	6.00	71.00	2.68
13	6.50	72.30	1.80
14	7.00	73.70	1.90
15	7.50	74.90	1.60
16	8.00	76.00	1.45
17	8.50	76.80	1.04
18	9.00	77.70	1.16
19	9.50	78.50	1.02
20	10.00	79.20	0.88
21	10.50	79.90	0.88
22	11.00	80.50	0.75

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(80.5 - 17.37807) (1)/(1) = +6.31E+01 g/L Carbon
 (80.5 - 17.37807) (1)/(1)(12) = +5.26E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

55- .0996g

3917

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03391 DUP

Date: 12/19/98

Time: 14:52:09

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.60	0.00
2	1.01	1.50	60.00
3	1.51	41.90	96.42
4	2.00	178.00	76.46
5	2.50	277.40	35.83
6	3.00	345.00	19.59
7	3.50	390.80	11.72
8	4.00	418.50	6.62
9	4.50	432.20	3.17
10	5.00	438.90	1.53
11	5.50	442.30	0.77
12	6.00	444.40	0.47
13	6.50	446.00	0.36
14	7.00	447.40	0.31
15	7.50	448.40	0.22
16	8.00	449.30	0.20
17	8.50	450.20	0.20
18	9.00	451.20	0.22
19	9.50	452.00	0.18
20	10.00	452.80	0.18
21	10.50	453.50	0.15
22	11.00	454.30	0.18

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(454.3 - 7.148691) (1) / (1) = +4.472E+02 g/L Carbon
 (454.3 - 7.148691) (1) / (1) (12) = +3.726E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

552 .0860g

3918

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T033912 DU

Date: 12/19/98

Time: 15:05:08

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 1.58 ug/minute C

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.30	53.85
3	1.51	6.30	79.37
4	2.00	20.40	69.12
5	2.50	34.70	41.21
6	3.00	45.00	22.89
7	3.50	51.70	12.96
8	4.00	55.90	7.51
9	4.50	58.60	4.61
10	5.00	61.10	4.09
11	5.50	62.90	2.86
12	6.00	64.40	2.33
13	6.50	65.80	2.13
14	7.00	67.00	1.79
15	7.50	68.10	1.62
16	8.00	69.20	1.59
17	8.50	70.00	1.14
18	9.00	70.90	1.27
19	9.50	71.80	1.25
20	10.00	72.50	0.97
21	10.50	73.30	1.09
22	11.00	73.80	0.68

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(73.8 - 17.37527) (1)/(1) = +5.64E+01 g/L Carbon
 (73.8 - 17.37527) (1)/(1)(12) = +4.70E+00 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS- 09609

3919

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03391 SPK Date: 12/19/98 Time: 15:18:50

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = .65 ug/minute CAnalyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.20	50.00
3	1.51	24.10	95.02
4	2.00	166.50	85.53
5	2.50	358.60	53.57
6	3.00	527.20	31.98
7	3.50	638.10	17.38
8	4.00	714.30	10.67
9	4.50	767.30	6.91
10	5.00	797.50	3.79
11	5.50	809.10	1.43
12	6.00	814.10	0.61
13	6.50	816.80	0.33
14	7.00	818.70	0.23
15	7.50	820.30	0.20
16	8.00	821.70	0.17
17	8.50	822.90	0.15
18	9.00	824.00	0.13
19	9.50	825.10	0.13
20	10.00	826.20	0.13
21	10.50	827.30	0.13
22	11.00	828.10	0.10

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(828.1 - 7.148016) (1)/(1)	=	+8.210E+02	g/L Carbon
(828.1 - 7.148016) (1)/(1) (12)	=	+6.841E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0897g + .500ml spk 25N12D

3920

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03391 SPK Date: 12/19/98 Time: 15:32:26

Sample Size = 1 uL Analyst : SL HOOD
Dil Factor = 1 Min Readings = 22
Blank ID # = Max Readings = 22
Blank Value = 1.58 ug/minute C % Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	1.00	0.00
2	1.01	2.00	50.00
3	1.51	21.60	90.74
4	2.01	95.30	77.33
5	2.51	188.80	49.52
6	3.01	261.00	27.66
7	3.51	303.60	14.03
8	4.00	324.60	6.47
9	4.50	333.70	2.73
10	5.00	338.60	1.45
11	5.50	341.90	0.97
12	6.00	344.30	0.70
13	6.50	346.20	0.55
14	7.00	348.20	0.57
15	7.50	349.60	0.40
16	8.00	351.00	0.40
17	8.50	352.30	0.37
18	9.00	353.40	0.31
19	9.50	354.40	0.28
20	10.00	355.50	0.31
21	10.50	356.30	0.22
22	11.00	357.10	0.22

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(357.1 - 17.37807) (1)/(1) = +3.397E+02 g/L Carbon
 (357.1 - 17.37807) (1)/(1) (12) = +2.831E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0897g + .100 ml spk 26M12-F
3921

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003396

Date: 12/19/98

Time: 15:45:00

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .65 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	0.90	0.00
2	1.01	1.50	40.00
3	1.50	3.10	51.61
4	2.00	12.70	75.59
5	2.50	35.30	64.02
6	3.00	66.20	46.68
7	3.50	95.60	30.75
8	4.00	116.00	17.59
9	4.50	127.80	9.23
10	5.00	133.90	4.56
11	5.50	137.40	2.55
12	6.00	139.80	1.72
13	6.50	141.60	1.27
14	7.00	143.30	1.19
15	7.50	144.70	0.97
16	8.00	145.90	0.82
17	8.50	147.20	0.88
18	9.00	148.40	0.81
19	9.50	149.60	0.80
20	10.00	150.80	0.80
21	10.50	151.60	0.53
22	11.00	152.80	0.79

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(152.8 - 7.149246) (1)/(1) =

+1.457E+02 g/L Carbon

(152.8 - 7.149246) (1)/(1)(12) =

+1.214E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .03393

3922

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T003396

Date: 12/19/98

Time: 15:58:36

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 1.58 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	1.10	0.00
2	1.00	2.20	50.00
3	1.50	26.00	91.54
4	2.00	101.40	74.36
5	2.50	199.70	49.22
6	3.00	299.30	33.28
7	3.50	387.10	22.68
8	4.00	461.00	16.03
9	4.50	523.50	11.94
10	5.00	574.80	8.92
11	5.50	615.80	6.66
12	6.00	646.80	4.79
13	6.50	670.50	3.53
14	7.00	689.10	2.70
15	7.50	704.20	2.14
16	8.00	716.90	1.77
17	8.50	727.40	1.44
18	9.00	736.20	1.20
19	9.50	743.50	0.98
20	10.00	749.70	0.83
21	10.50	754.90	0.69
22	11.00	759.40	0.59

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(759.4 - 17.37663) (1)/(1) = +7.420E+02 g/L Carbon
 (759.4 - 17.37663) (1)/(1)(12) = +6.184E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0339g

3923

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03396 DUP Date: 12/19/98 Time: 16:10:49

Sample Size = 1 uL Analyst : SL HOOD
 Dil Factor = 1 Min Readings = 22
 Blank ID # = Max Readings = 22
 Blank Value = .65 ug/minute C % Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	2.60	0.00
2	1.01	4.70	44.68
3	1.51	7.80	39.74
4	2.01	21.70	64.06
5	2.51	48.60	55.35
6	3.01	81.80	40.59
7	3.51	110.40	25.91
8	4.00	128.70	14.22
9	4.50	138.50	7.08
10	5.00	143.70	3.62
11	5.50	146.70	2.04
12	6.00	148.90	1.48
13	6.50	150.70	1.19
14	7.00	152.30	1.05
15	7.50	153.80	0.98
16	8.00	155.10	0.84
17	8.50	156.40	0.83
18	9.00	157.60	0.76
19	9.50	158.80	0.76
20	10.00	159.90	0.69
21	10.50	161.10	0.74
22	11.00	162.10	0.62

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(162.1 - 7.150436) (1)/(1) = +1.549E+02 g/L Carbon
 (162.1 - 7.150436) (1)/(1):(12) = +1.291E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

3924

SS = .0304g

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03396 DUP Date: 12/19/98 Time: 16:24:16

Sample Size = 1 uL Analyst : SL HOOD
Dil Factor = 1 Min Readings = 22
Blank ID # = Max Readings = 22
Blank Value = 1.58 ug/minute C % Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.51	1.10	0.00
2	1.01	2.40	54.17
3	1.50	23.20	89.66
4	2.00	88.60	73.81
5	2.50	172.90	48.76
6	3.00	259.60	33.40
7	3.50	335.90	22.72
8	4.00	400.70	16.17
9	4.50	453.90	11.72
10	5.00	497.10	8.69
11	5.50	530.30	6.26
12	6.00	555.80	4.59
13	6.50	575.70	3.46
14	7.00	591.60	2.69
15	7.50	604.90	2.20
16	8.00	616.00	1.80
17	8.50	624.80	1.41
18	9.00	632.40	1.20
19	9.50	638.30	0.92
20	10.00	643.40	0.79
21	10.50	647.80	0.68
22	11.00	651.30	0.54

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(651.3 - 17.37527) (1)/(1) = +6.339E+02 g/L Carbon
 (651.3 - 17.37527) (1)/(1) (12) = +5.283E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0304g

3925

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03396 SPK Date: 12/19/98 Time: 16:38:36

Sample Size = 1 uL Analyst : SL HOOD
 Dil Factor = 1 Min Readings = 22
 Blank ID # = Max Readings = 22
 Blank Value = .65 ug/minute C % Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.51	1.20	0.00
2	1.01	2.30	47.83
3	1.50	13.50	82.96
4	2.00	85.60	84.23
5	2.50	183.20	53.28
6	3.00	268.40	31.74
7	3.50	333.90	19.62
8	4.00	378.90	11.88
9	4.50	403.30	6.05
10	5.00	414.10	2.61
11	5.50	419.40	1.26
12	6.00	422.70	0.78
13	6.50	425.00	0.54
14	7.00	427.10	0.49
15	7.50	428.90	0.42
16	8.00	430.50	0.37
17	8.50	432.10	0.37
18	9.00	433.50	0.32
19	9.50	434.80	0.30
20	10.00	436.20	0.32
21	10.50	437.30	0.25
22	11.00	438.50	0.27

USER INPUT BLANK VALUE

BLANK VALUE = 7.148691 micrograms carbon

BLANK FACTOR = 7.148691 / 10.99799 = +6.5E-01 ug/min Carbon

SAMPLE RESULTS:

(438.5 - 7.148056) (1)/(1) = +4.314E+02 g/L Carbon
 (438.5 - 7.148056) (1)/(1)(12) = +3.595E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .0284g + .500ml spk 25N12-D
 3926

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T03396 SPK Date: 12/19/98 Time: 16:55:09

Sample Size = 1 uL Analyst : SL HOOD
Dil Factor = 1 Min Readings = 22
Blank ID # = Max Readings = 22
Blank Value = 1.58 ug/minute C % Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference
1	0.51	1.30	0.00
2	1.01	3.00	56.67
3	1.51	36.70	91.83
4	2.00	125.00	70.64
5	2.50	257.70	51.49
6	3.00	412.30	37.50
7	3.50	533.90	22.78
8	4.00	617.00	13.47
9	4.50	679.10	9.14
10	5.00	725.40	6.38
11	5.50	757.80	4.28
12	6.00	779.90	2.83
13	6.50	795.90	2.01
14	7.00	808.60	1.57
15	7.50	819.60	1.34
16	8.00	828.80	1.11
17	8.50	836.60	0.93
18	9.00	843.20	0.78
19	9.50	849.10	0.69
20	10.00	854.50	0.63
21	10.50	859.00	0.52
22	11.00	862.90	0.45

USER INPUT BLANK VALUE

BLANK VALUE = 17.37682 micrograms carbon

BLANK FACTOR = 17.37682 / 10.99799 = +1.6E+00 ug/min Carbon

SAMPLE RESULTS:

(862.9 - 17.37682) (1)/(1) = +8.455E+02 g/L Carbon
(862.9 - 17.37682) (1)/(1) (12) = +7.046E+01 Molar Carbon
<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .0284g + .100 ml spk 26N12-F
3927

LABCORE Completed Worklist Report for Worklist# 28061

Analyst: pjm

Instrument: CARB2

Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY105 FIELD BLANK FOR @TICTOC1 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@TICTOC1 TIC-02	LIQUID	6.02E+02	6.08E+2	100.997 % Recovery	
1 STD	0		@TICTOC1 TOC-02	LIQUID	3.00E+03	2.86E+3	95.333 % Recovery	
2 BLNK	0		@TICTOC1 TIC-02	LIQUID	1	1.00E-1	0.100 ug/mL	
2 BLNK	0		@TICTOC1 TOC-02	LIQUID	1	4.00E+0	4.000 ug/mL	
3 SAMPLE	S98T002391	0	@TICTOC1 TIC-02	LIQUID	N/A	5.00E+00	5.000 ug/mL	
3 SAMPLE	S98T002391	0	@TICTOC1 TOC-02	LIQUID	N/A	4.00E+01	40.000 ug/mL	
4 DUP	S98T002391	0	@TICTOC1 TIC-02	LIQUID	<5.00E+0	<5.00E+0	RFD	
4 DUP	S98T002391	0	@TICTOC1 TOC-02	LIQUID	<4.00E+1	<4.00E+1	RFD	

Final page for worklist# 28061

Analyst Signature _____

Date _____

Mary Shan 1-29-99
Analyst Signature _____ Date _____

[Signature] 1/29/99
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 28061

Analyst: R. J. McCann Instrument: CARB2 _____ Book# TIC 25N12D

Method: LA-342-100 Rev/Mod F-1 TOC 34N12A

Worklist Comment: BY105 FIELD BLANK FOR @TICTOC1 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@TICTOC1	LIQUID		
2 BLNK			@TICTOC1	LIQUID		
3 SAMPLE	S98T002391 0		@TICTOC1	LIQUID	98000436	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002391 0		@TICTOC1	LIQUID		

Final page for worklist # 28061

R. J. McCann 1/27/99
Analyst Signature Date

Mary Fran 1-29-99
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	0.2000
STD	Dilution Factor (DF)	1	1
Work List	Final Coulometer Reading in μg (C1)	613.6	592.8
28061	μg of Carbon from Baseline (C2)	5.6	21
Test Code	Standard Book Number	25N12D	34N12A
@TICTOC1	Standard Value ($\mu\text{g}/\text{mL}$)	602	3000
Matrix	QC Actual in $\mu\text{g}/\text{mL}$ = Standard Value ($\mu\text{g}/\text{mL}$) QC Found in $\mu\text{g}/\text{mL}$ = $(C1 - C2) * DF / SS$ QC Found in $\mu\text{g}/\text{mL}$ for TIC = 5 if $C1 < C2$ QC Found in $\mu\text{g}/\text{mL}$ for TOC = 40 if $C1 < C2$ % Recovery = QC Found / QC Actual * 100		
LIQUID			
Batch Number			
99000333			
Rerun			
0			
Sample Prep			
N/A			
Sample #			
STD			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
PJM			
Date Completed			
01/29/99			
Analysis Date			
01/27/99			
Analysis Time	Method Detection Limit in $\mu\text{g}/\text{mL}$	5	40
11:50 PM	QC Actual in $\mu\text{g}/\text{mL}$	6.02E+02	3.00E+03
Sample Point	QC Found in $\mu\text{g}/\text{mL}$	6.08E+02	2.86E+03
BY 105	Percent Standard Recovery	101.0	95.3

Data Entered By:	MF	Date:	01/29/99
Signature of Chemist:	<i>NA</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3930

WORKBOOK PAGE: BLANK2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Worklist	µg of Carbon in Sample (C1)	5.5	17
28061	µg of Carbon from Baseline (C2)	5.6	21
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000333			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
01/29/99			
Analysis Date			
01/27/99			
Analysis Time	Method Detection Limit in µg/mL	5	40
11:50 PM			
Sample Point	µg of Carbon	1.00E-01	4.00E+00
BY 105			

Data Entered By:	MF	Date:	01/29/99
Signature of Chemist:	<i>MF</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3931

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	1.0000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	9.9	16.9
28061	µg of Carbon from Baseline (C2)	5.6	21
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000333			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002391			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
01/29/99	NOTE: FOR TIC: The Reported Result is < 5.		
Analysis Date	NOTE: FOR TOC: The Reported Result is < 40.		
01/27/99			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:50 PM		5	40
Sample Point	µg of Carbon/mL Note: <	4.30E+00	4.00E+01
BY 105			

Data Entered By:

MF

Date: 01/29/99

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	1.0000	1.0000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	9.6	20.3
28061	µg of Carbon from Baseline (C2)	5.6	21
Test Code	Known µg of C from Original Sample	<4.30E+0	<4.00E+1
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000333			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002391			
Instrument Code			
CARB2			
Prepared By			
MF			
Chemist			
MJL			
Analyst			
PJM			
Date Complete			
01/29/99	NOTE: FOR TIC: The Reported Result is < 5.		
Analysis Date	NOTE: FOR TOC: The Reported Result is < 40.		
01/27/99			
Analysis Time	Method Detection Limit in µg/mL	TIC	TOC
11:50 PM		5	40
Sample Point	µg of Carbon/mL Note: <	4.00E+00	4.00E+01
BY 105			

Data Entered By:	MF	Date:	01/29/99
Signature of Chemist:	N/A	Date:	

SAMPLE.WB1 REV 1.0

342100ML

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 01/27/99

Time: 22:40:38

Sample Size = 1000 uL
Dil Factor = 1
Blank ID # =
Blank Value = .53 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.08	0.00	0.00
2	0.51	0.80	100.00
3	1.00	8.50	90.59
4	1.51	129.00	93.41
5	2.00	321.20	59.84
6	2.50	468.50	31.44
7	3.00	549.30	14.71
8	3.50	588.10	6.60
9	4.00	602.60	2.41
10	4.50	606.60	0.66
11	5.00	608.50	0.31
12	5.50	609.60	0.18
13	6.00	610.10	0.08
14	6.50	610.60	0.08
15	7.00	611.00	0.07
16	7.50	611.40	0.07
17	8.00	611.90	0.08
18	8.50	612.30	0.07
19	9.00	612.60	0.05
20	9.50	612.90	0.05
21	10.00	613.30	0.07
22	10.50	613.60	0.05

USER INPUT BLANK VALUE

BLANK VALUE = 5.564482 micrograms carbon

BLANK FACTOR = 5.564482 / 10.49902 = +5.3E-01 ug/min Carbon

SAMPLE RESULTS:

(613.6 - 5.564029) (1) / (1000) = +6.080E-01 g/L Carbon
(613.6 - 5.564029) (1) / (1000) (12) = +5.067E-02 Molar Carbon

SIGNATURE BELOW REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3934-3945

Sample Run By: PJ McCown 1/27/99
PJ MCCOWN 00000

1.0 mL 25N12D

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD-2

Date: 01/27/99

Time: 22:53:43

Sample Size = 200 uL
Dil Factor = 1
Blank ID # =
Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.50	0.20	100.00
3	1.01	1.00	80.00
4	1.50	7.40	86.49
5	2.00	42.50	82.59
6	2.50	136.60	68.89
7	3.00	304.50	55.14
8	3.50	463.80	34.35
9	4.00	541.60	14.36
10	4.50	572.30	5.36
11	5.00	581.70	1.62
12	5.50	585.20	0.60
13	6.00	586.90	0.29
14	6.50	587.70	0.14
15	7.00	588.60	0.15
16	7.50	589.60	0.17
17	8.00	590.30	0.12
18	8.50	590.80	0.08
19	9.00	591.40	0.10
20	9.50	591.90	0.08
21	10.00	592.40	0.08
22	10.50	592.80	0.07

USER INPUT BLANK VALUE

BLANK VALUE = 20.99805 micrograms carbon

BLANK FACTOR = 20.99805 / 10.49902 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(592.8 - 20.99438) (1) / (200) = +2.859E+00 g/L Carbon
 (592.8 - 20.99438) (1) / (200) (12) = +2.383E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ McCown 1/27/99
PJ MCCOWN 00000

-200mL 34 N/2A

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0
<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 01/27/99

Time: 22:13:21

Sample Size = 1 uL
Dil Factor = 1
Blank ID # = BASE-2
Blank Value = N/A

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	0.50	60.00
4	1.51	0.80	37.50
5	2.00	1.30	38.46
6	2.50	1.60	18.75
7	3.00	1.90	15.79
8	3.50	2.10	9.52
9	4.00	2.40	12.50
10	4.50	2.60	7.69
11	5.00	2.90	10.34
12	5.50	3.10	6.45
13	6.00	3.40	8.82
14	6.50	3.60	5.56
15	7.00	3.80	5.26
16	7.50	4.10	7.32
17	8.00	4.40	6.82
18	8.50	4.60	4.35
19	9.00	4.90	6.12
20	9.50	5.10	3.92
21	10.00	5.30	3.77
22	10.50	5.60	5.36

BLANK VALUE = 5.6 micrograms carbon

BLANK FACTOR = 5.6 / 10.49902 =

+5.3E-01 ug/min Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0
<<< BLANK ANALYSIS >>>

Sample: BASE-2

Date: 01/27/99

Time: 22:26:07

Sample Size = 1 uL
Dil Factor = 1
Blank ID # = BASE-2
Blank Value = N/A

Analyst : PJ MCCOWN
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	Analysis	Time	==== Coulometer ==	% Difference ==
1	0.08		0.00	0.00
2	0.51		0.30	100.00
3	1.01		1.20	75.00
4	1.51		3.30	63.64
5	2.00		6.40	48.44
6	2.50		9.20	30.43
7	3.00		11.60	20.69
8	3.50		12.80	9.37
9	4.00		13.70	6.57
10	4.50		14.50	5.52
11	5.00		15.10	3.97
12	5.50		15.50	2.58
13	6.00		15.90	2.52
14	6.50		16.50	3.64
15	7.00		17.20	4.07
16	7.50		17.80	3.37
17	8.00		18.40	3.26
18	8.50		18.90	2.65
19	9.00		19.50	3.08
20	9.50		20.00	2.50
21	10.00		20.50	2.44
22	10.50		21.00	2.38

BLANK VALUE = 21 micrograms carbon

BLANK FACTOR = 21 / 10.49902 =

+2.00E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 01/27/99

Time: 23:12:22

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .53 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	==== Coulometer	==== % Difference ==
1	0.08	0.00	0.00
2	0.50	0.10	100.00
3	1.00	0.40	75.00
4	1.51	0.70	42.86
5	2.00	1.10	36.36
6	2.50	1.30	15.38
7	3.00	1.50	13.33
8	3.50	1.90	21.05
9	4.00	2.10	9.52
10	4.50	2.40	12.50
11	5.00	2.60	7.69
12	5.50	3.00	13.33
13	6.00	3.20	6.25
14	6.50	3.40	5.88
15	7.00	3.60	5.56
16	7.50	3.90	7.69
17	8.00	4.20	7.14
18	8.50	4.50	6.67
19	9.00	4.80	6.25
20	9.50	5.00	4.00
21	10.00	5.20	3.85
22	10.50	5.50	5.45

USER INPUT BLANK VALUE

BLANK VALUE = 5.564482 micrograms carbon

BLANK FACTOR = 5.564482 / 10.49902 = +5.3E-01 ug/min Carbon

SAMPLE RESULTS:

(5.5 - 5.563964) (1)/(1) =

< 5.00 E-3 g/L Carbon

(5.5 - 5.563964) (1)/(1)(12) =

< 4.17 E-4 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLNK-2

Date: 01/27/99

Time: 23:25:02

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.10	100.00
3	1.01	1.00	90.00
4	1.51	3.00	66.67
5	2.00	5.90	49.15
6	2.50	8.00	26.25
7	3.00	9.80	18.37
8	3.50	10.70	8.41
9	4.00	11.40	6.14
10	4.50	12.00	5.00
11	5.00	12.70	5.51
12	5.50	13.20	3.79
13	6.00	13.60	2.94
14	6.50	14.20	4.23
15	7.00	14.50	2.07
16	7.50	15.10	3.97
17	8.00	15.30	1.31
18	8.50	15.70	2.55
19	9.00	16.00	1.88
20	9.50	16.40	2.44
21	10.00	16.70	1.80
22	10.50	17.00	1.76

USER INPUT BLANK VALUE

BLANK VALUE = 20.99805 micrograms carbon

BLANK FACTOR = 20.99805 / 10.49902 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(17 - 20.99805) (1)/(1) = < 5.00 E-3 g/L Carbon
 (17 - 20.99805) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

PJ MCCOWN

00000

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002391

Date: 01/27/99

Time: 23:37:26

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .53 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

Reading	Analysis Time	Coulometer	% Difference
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	1.90	84.21
4	1.51	3.40	44.12
5	2.00	4.50	24.44
6	2.50	5.20	13.46
7	3.00	5.60	7.14
8	3.50	6.10	8.20
9	4.00	6.40	4.69
10	4.50	6.70	4.48
11	5.00	6.80	1.47
12	5.50	7.20	5.56
13	6.00	7.40	2.70
14	6.50	7.70	3.90
15	7.00	7.90	2.53
16	7.50	8.30	4.82
17	8.00	8.60	3.49
18	8.50	8.70	1.15
19	9.00	9.00	3.33
20	9.50	9.30	3.23
21	10.00	9.60	3.13
22	10.50	9.90	3.03

USER INPUT BLANK VALUE

BLANK VALUE = 5.564482 micrograms carbon

BLANK FACTOR = 5.564482 / 10.49902 = +5.3E-01 ug/min Carbon

SAMPLE RESULTS:

(9.9 - 5.564547) (1)/(1) =

+4.3E+00 g/L Carbon

(9.9 - 5.564547) (1)/(1) (12) =

+3.6E-01 Molar Carbon

Sample Run By:

PJ MCCOWN

00000

1.0 uL C

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002391

Date: 01/27/99

Time: 23:49:47

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.08	0.00	0.00
2	0.51	0.40	100.00
3	1.01	1.30	69.23
4	1.51	2.60	50.00
5	2.00	4.90	46.94
6	2.50	7.50	34.67
7	3.00	9.60	21.88
8	3.50	10.70	10.28
9	4.00	11.80	9.32
10	4.50	12.40	4.84
11	5.00	12.90	3.88
12	5.50	13.30	3.01
13	6.00	13.80	3.62
14	6.50	14.20	2.82
15	7.00	14.50	2.07
16	7.50	15.00	3.33
17	8.00	15.30	1.96
18	8.50	15.70	2.55
19	9.03	15.90	1.26
20	9.53	16.30	2.45
21	10.03	16.60	1.81
22	10.53	16.90	1.78

USER INPUT BLANK VALUE

BLANK VALUE = 20.99805 micrograms carbon

BLANK FACTOR = 20.99805 / 10.49902 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(16.9 - 21.06592) (1) / (1) = < 5.00 E-3 g/L Carbon
 (16.9 - 21.06592) (1) / (1) (12) = < 4.17 E-4 Molar Carbon
 <<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>

Sample Run By:

PJ MCCOWN

00000

1.0 M C

3941

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2391 DUP

Date: 01/28/99

Time: 00:01:38

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .53 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time ==	Coulometer ==	% Difference ==
1	0.08	0.00	0.00
2	0.51	0.30	100.00
3	1.01	1.10	72.73
4	1.50	2.80	60.71
5	2.00	4.00	30.00
6	2.50	4.90	18.37
7	3.00	5.50	10.91
8	3.50	5.90	6.78
9	4.00	6.20	4.84
10	4.50	6.50	4.62
11	5.00	6.80	4.41
12	5.50	7.00	2.86
13	6.00	7.20	2.78
14	6.50	7.60	5.26
15	7.00	7.80	2.56
16	7.50	8.00	2.50
17	8.00	8.30	3.61
18	8.50	8.60	3.49
19	%-1431.00	8.90	3.37
20	%-1430.50	9.00	1.11
21	%-1430.00	9.30	3.23
22	%-1429.50	9.60	3.13

USER INPUT BLANK VALUE

BLANK VALUE = 5.564482 micrograms carbon

BLANK FACTOR = 5.564482 / 10.49902 = +5.3E-01 ug/min Carbon

SAMPLE RESULTS:

(9.6 --757.636) (1) / (1) = +7.7E+02 g/L Carbon
 (9.6 --757.636) (1) / (1) (12) = +6.4E+01 Molar Carbon

Sample Run By: PJ MCCOWN

PJ MCCOWN

00000

1.0 mL

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: 2391 DUP

Date: 01/28/99

Time: 02:04:11

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2 ug/minute C

Analyst : PJ MCCOWN

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference
1	0.08	0.00	0.00
2	0.51	0.20	100.00
3	1.01	1.20	83.33
4	1.51	3.70	67.57
5	2.00	6.60	43.94
6	2.50	8.90	25.84
7	3.00	11.10	19.82
8	3.50	12.40	10.48
9	4.00	13.50	8.15
10	4.50	14.40	6.25
11	5.00	15.10	4.64
12	5.50	15.60	3.21
13	6.00	16.30	4.29
14	6.50	16.80	2.98
15	7.00	17.40	3.45
16	7.50	17.80	2.25
17	8.00	18.20	2.20
18	8.50	18.70	2.67
19	9.00	19.10	2.09
20	9.50	19.60	2.55
21	10.00	20.00	2.00
22	10.50	20.30	1.48

USER INPUT BLANK VALUE

BLANK VALUE = 20.99805 micrograms carbon

BLANK FACTOR = 20.99805 / 10.49902 = +2.0E+00 ug/min Carbon

SAMPLE RESULTS:

(20.3 - 20.99633) (1)/(1) = < 5.00 E-3 g/L Carbon
 (20.3 - 20.99633) (1)/(1) (12) = < 4.17 E-4 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By: PJ MCCOWN

PJ MCCOWN

00000

1.0 uL

LABCORE Completed Worklist Report for Worklist# 28635

Analyst: slh Instrument: CARB2 Book#: _____

Method: LA-342-100 Rev/Mod _____

Worklist Comment: BY - 105 RERUN jmv

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 BLNK	0		@TICTOCI TIC-02	LIQUID	1	1.00E-0	1.000	ug/mL
1 BLNK	0		@TICTOCI TOC-02	LIQUID	1	1.30E+1	13.000	ug/mL
2 STD	0		@TICTOCI TIC-02	LIQUID	6.02E+02	6.09E+2	101.163	% Recovery
2 STD	0		@TICTOCI TOC-02	LIQUID	3.00E+03	2.80E+3	93.333	% Recovery
3 SAMPLE	S98T002967	0	@TICTOCI TIC-02	LIQUID	N/A	5.62E+03	5.000	ug/mL
3 SAMPLE	S98T002967	0	@TICTOCI TOC-02	LIQUID	N/A	2.68E+03	40.000	ug/mL
4 DUP	S98T002967	0	@TICTOCI TIC-02	LIQUID	5.62E+3	5.59E+3	0.535	RPD
4 DUP	S98T002967	0	@TICTOCI TOC-02	LIQUID	2.68E+3	2.72E+3	1.481	RPD
5 SAMPLE	S98T002968	0	@TICTOCI TIC-02	LIQUID	N/A	3.71E+03	5.000	ug/mL
5 SAMPLE	S98T002968	0	@TICTOCI TOC-02	LIQUID	N/A	2.62E+03	40.000	ug/mL
6 DUP	S98T002968	0	@TICTOCI TIC-02	LIQUID	3.71E+3	3.58E+3	3.567	RPD
6 DUP	S98T002968	0	@TICTOCI TOC-02	LIQUID	2.62E+3	2.51E+3	4.288	RPD

Final page for worklist# 28635

Analyst Signature _____ Date _____

3-15-99
Analyst Signature _____ Date _____

Reviewer Signature _____ Date 3/15/99

LABCORE Data Entry Template for Worklist# 28635

Analyst: JH Instrument: CARB2 Book# 25ND-D
Method: LA-342-100 Rev/Mod F-2 34N/2-A
Worklist Comment: BY - 105 RERUN jmv

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 BLNK			@TICTOC1	LIQUID		
2 STD			@TICTOC1	LIQUID		
3 SAMPLE	S98T002967 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
4 DUP	S98T002967 0		@TICTOC1	LIQUID		
5 SAMPLE	S98T002968 0		@TICTOC1	LIQUID	98000509	BY-105 (2)
	Analytes Requested: TIC-02 , TOC-02					
6 DUP	S98T002968 0		@TICTOC1	LIQUID		

Final page for worklist # 28635

J. Hood Boatright
Signature Date
3-12-99

Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

3945

WORKBOOK PAGE: BLANK1

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.0000	0.0000
BLNK	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	8.7	17.9
28635	µg of Carbon from Baseline (C2)	7.7	30.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000952			
Rerun	µg of Carbon = C1-C2		
0			
Sample Prep			
N/A			
Sample #			
BLNK			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
03/15/99			
Analysis Date			
03/12/99			
Analysis Time	Method Detection Limit in µg/mL	5	40
01:10 PM			
Sample Point	µg of Carbon	1.00E+00	1.30E+01
BY 105			

Data Entered By:	JMV	Date:	03/15/99
Signature of Chemist:	<i>[Signature]</i>	Date:	

BLANK.WB1 REV 1.0

342100ML

3946

WORKBOOK PAGE: STD2

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

			TIC	TOC
Type	Sample Size in mL	(SS)	1.0000	0.2000
STD	Dilution Factor	(DF)	1	1
Work List	Final Coulometer Reading in µg	(C1)	616.3	590.6
28635	µg of Carbon from Baseline	(C2)	7.7	30.9
Test Code	Standard Book Number		25N12D	34N12A
@TICTOC1	Standard Value (µg/ml)		602	3000

Matrix
LIQUIDBatch Number
99000952Rerun
0Sample Prep
N/ASample #
STDInstrument Code
CARB2Prepared By
JMVChemist
MJLAnalyst
SLHDate Complete
03/15/99Analysis Date
03/12/99Analysis Time
01:10 PMSample Point
BY 105

QC Actual in µg/mL = Standard Value (µg/mL)

QC Found in µg/mL = (C1 - C2) * DF / SS

QC Found in µg/mL for TIC = 5 if C1 < C2

QC Found in µg/mL for TOC = 40 if C1 < C2

% Recovery = QC Found / QC Actual * 100

		TIC	TOC
Analysis Time	Method Detection Limit in µg/mL	5	40
01:10 PM	QC Actual in µg/mL	6.02E+02	3.00E+03
Sample Point	QC Found in µg/mL	6.09E+02	2.80E+03
BY 105	Percent Standard Recovery	101.1	93.3

Data Entered By:	JMV	Date:	03/15/99
Signature of Chemist:	<i>[Signature]</i>	Date:	

STANDARD.WB1 REV 1.0

342100ML

3947

WORKBOOK PAGE: SAM3

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	570.1	298.7
28635	µg of Carbon from Baseline (C2)	7.7	30.9
Test Code			
@TICTOC1			

Matrix
LIQUID

Batch Number
99000952

Rerun
0

µg of Carbon/mL = (C1-C2) * DF / SS
 µg of Carbon/mL for TIC = 5 if C1 < C2
 µg of Carbon/mL for TOC = 40 if C1 < C2

Sample Prep
N/A

Sample #
S98T002967

Instrument Code
CARB2

Prepared By
JMV

Chemist
MJL

Analyst
SLH

Date Complete
03/15/99

Analysis Date
03/12/99

Analysis Time
01:10 PM

Method Detection Limit in µg/mL

Sample Point
BY 105

µg of Carbon/mL

TIC	TOC
5	40
5.62E+03	2.68E+03

Data Entered By:

JMV

Date: 03/15/99

Signature of Chemist:

Date:

WORKBOOK PAGE: DUP4

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	566.2	302.9
28635	µg of Carbon from Baseline (C2)	7.7	30.9
Test Code	Known µg of C from Original Sample	5.62E+3	2.68E+3
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000952			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002967			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
03/15/99			
Analysis Date			
03/12/99			
Analysis Time	Method Detection Limit in µg/mL	5	40
01:10 PM			
Sample Point	µg of Carbon/mL	5.59E+03	2.72E+03
BY 105			

Data Entered By: _____ JM V Date: 03/15/99

Signature of Chemist: Date: _____

SAMPLE.WB1 REV 1.0

342100M*

3949

WORKBOOK PAGE: SAM5

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
SAMPLE	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	378.7	293.2
28635	µg of Carbon from Baseline (C2)	7.7	30.9
Test Code			
@TICTOC1			
Matrix			
LIQUID			
Batch Number			
99000952			
Rerun	µg of Carbon/mL = (C1-C2) * DF / SS		
0	µg of Carbon/mL for TIC = 5 if C1 < C2		
Sample Prep	µg of Carbon/mL for TOC = 40 if C1 < C2		
N/A			
Sample #			
S98T002968			
Instrument Code			
CARB2			
Prepared By			
JMV			
Chemist			
MJL			
Analyst			
SLH			
Date Complete			
03/15/99			
Analysis Date			
03/12/99			
Analysis Time	Method Detection Limit in µg/mL	5	40
01:10 PM			
Sample Point	µg of Carbon/mL	3.71E+03	2.62E+03
BY 105			

Data Entered By:

JMV

Date:

03/15/99

Signature of Chemist:

NA

Date:

WORKBOOK PAGE: DUP6

TIC/TOC : LA-342-100 (F-1)

LIQUIDS

		TIC	TOC
Type	Sample Size in mL (SS)	0.1000	0.1000
DUP	Dilution Factor (DF)	1	1
Work List	µg of Carbon in Sample (C1)	365.6	281.5
28635	µg of Carbon from Baseline (C2)	7.7	30.9
Test Code	Known µg of C from Original Sample	3.71E+3	2.62E+3
@TICTOC1			

Matrix
LIQUID

Batch Number
99000952

Rerun
0

µg of Carbon/mL = (C1-C2) * DF / SS
µg of Carbon/mL for TIC = 5 if C1 < C2
µg of Carbon/mL for TOC = 40 if C1 < C2

Sample Prep
N/A

Sample #
S98T002968

Instrument Code
CARB2

Prepared By
JMV

Chemist
MJL

Analyst
SLH

Date Complete
03/15/99

Analysis Date
03/12/99

Analysis Time
01:10 PM

Method Detection Limit in µg/mL

Sample Point
BY 105

µg of Carbon/mL

TIC	TOC
5	40
3.58E+03	2.51E+03

Data Entered By:

JMV

Date: 03/15/99

Signature of Chemist:

Date:

SAMPLE.WB1 REV 1.0

342100ML

3951

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 03/12/99

Time: 09:52:32

Sample Size = 1000 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	1.00	70.00
3	1.50	39.70	97.48
4	2.00	211.70	81.25
5	2.50	383.00	44.73
6	3.00	494.80	22.59
7	3.50	557.90	11.31
8	4.00	589.20	5.31
9	4.50	602.50	2.21
10	5.00	607.90	0.89
11	5.50	610.10	0.36
12	6.00	611.30	0.20
13	6.50	612.10	0.13
14	7.00	612.70	0.10
15	7.50	613.30	0.10
16	8.00	613.80	0.08
17	8.50	614.30	0.08
18	9.00	614.70	0.07
19	9.50	615.10	0.07
20	10.00	615.60	0.08
21	10.50	615.90	0.05
22	11.00	616.30	0.06

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(616.3 - 7.697308) (1) / (1000) = +6.086E-01 g/L Carbon
 (616.3 - 7.697308) (1) / (1000) (12) = +5.072E-02 Molar Carbon

Sample Run By:

SL HOOD

00002

SIGNATURE ABOVE REPRESENTS CHEMICAL TECHNOLOGIST/CHEMIST THAT
 COMPLETED/VERIFIED THE CALIBRATION/ANALYSIS ON PAGES 3954 TO 3967.

3952

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: STD 2

Date: 03/12/99

Time: 10:17:36

Sample Size = 200 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.81 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.30	0.00
2	1.01	1.20	75.00
3	1.51	42.30	97.16
4	2.01	170.70	75.22
5	2.51	343.20	50.26
6	3.01	473.40	27.50
7	3.51	538.40	12.07
8	4.01	565.30	4.76
9	4.50	575.20	1.72
10	5.00	579.50	0.74
11	5.50	581.50	0.34
12	6.00	583.20	0.29
13	6.50	584.20	0.17
14	7.00	585.10	0.15
15	7.50	586.00	0.15
16	8.00	586.70	0.12
17	8.50	587.50	0.14
18	9.00	588.10	0.10
19	9.50	588.60	0.08
20	10.00	589.20	0.10
21	10.50	589.90	0.12
22	11.00	590.60	0.12

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(590.6 - 30.90691) (1)/(200) = +2.798E+00 g/L Carbon
 (590.6 - 30.90691) (1)/(200) (12) = +2.332E-01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL Hood
 SL HOOD

00002

35 .200ml = 34N12-A

3953

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 03/12/99

Time: 09:28:19

Sample Size = 1 uL
 Dil Factor = 1
 Blank ID # = BASE 2
 Blank Value = N/A

Analyst : SL HOOD
 Min Readings = 22
 Max Readings = 22
 % Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.90	0.00
2	1.00	1.40	35.71
3	1.50	1.70	17.65
4	2.00	2.10	19.05
5	2.50	2.70	22.22
6	3.00	3.00	10.00
7	3.50	3.30	9.09
8	4.00	3.60	8.33
9	4.50	4.00	10.00
10	5.00	4.30	6.98
11	5.50	4.50	4.44
12	6.00	4.90	8.16
13	6.50	5.10	3.92
14	7.00	5.40	5.56
15	7.50	5.70	5.26
16	8.00	6.00	5.00
17	8.50	6.30	4.76
18	9.00	6.50	3.08
19	9.50	6.90	5.80
20	10.00	7.10	2.82
21	10.50	7.40	4.05
22	11.00	7.70	3.90

BLANK VALUE = 7.7 micrograms carbon

BLANK FACTOR = 7.7 / 10.99878 =

+7.0E-01

ug/min Carbon

Sample Run By:

SL HOOD

00002

3954

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT

TICTOC REV 2.0

<<< BLANK ANALYSIS >>>

Sample: BASE 2

Date: 03/12/99

Time: 09:40:18

Sample Size = 1 uL

Dil Factor = 1

Blank ID # = BASE 2

Blank Value = N/A

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	0.80	62.50
3	1.51	6.50	87.69
4	2.00	13.40	51.49
5	2.50	18.30	26.78
6	3.00	21.60	15.28
7	3.50	23.60	8.47
8	4.00	24.90	5.22
9	4.50	25.70	3.11
10	5.00	26.40	2.65
11	5.50	27.00	2.22
12	6.00	27.40	1.46
13	6.50	27.90	1.79
14	7.00	28.30	1.41
15	7.50	28.70	1.39
16	8.00	29.00	1.03
17	8.50	29.40	1.36
18	9.00	29.70	1.01
19	9.50	30.00	1.00
20	10.00	30.30	0.99
21	10.50	30.60	0.98
22	11.00	30.90	0.97

BLANK VALUE = 30.9 micrograms carbon

BLANK FACTOR = 30.9 / 10.99805 =

+2.81E+00 ug/min Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3955

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 03/12/99

Time: 10:29:47

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	Analysis Time	Coulometer	% Difference ==
1	0.51	0.30	0.00
2	1.01	0.70	57.14
3	1.50	2.60	73.08
4	2.00	3.70	29.73
5	2.50	4.30	13.95
6	3.00	4.70	8.51
7	3.50	5.10	7.84
8	4.00	5.40	5.56
9	4.50	5.60	3.57
10	5.00	5.80	3.45
11	5.50	6.10	4.92
12	6.00	6.30	3.17
13	6.50	6.60	4.55
14	7.00	6.80	2.94
15	7.50	7.00	2.86
16	8.00	7.30	4.11
17	8.50	7.50	2.67
18	9.00	7.80	3.85
19	9.50	8.00	2.50
20	10.00	8.30	3.61
21	10.50	8.50	2.35
22	11.00	8.70	2.30

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(8.7 - 7.697308) (1)/(1)	=	+1.0E+00	g/L Carbon
(8.7 - 7.697308) (1)/(1) (12)	=	+8.4E-02	Molar Carbon

Sample Run By:

SL HOOD

00002

3956

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: BLK 2

Date: 03/12/99

Time: 10:46:53

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.81 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.30	0.00
2	1.01	0.50	40.00
3	1.50	2.60	80.77
4	2.00	5.70	54.39
5	2.50	8.30	31.33
6	3.00	10.40	20.19
7	3.50	11.70	11.11
8	4.00	12.80	8.59
9	4.50	13.40	4.48
10	5.00	14.00	4.29
11	5.50	14.40	2.78
12	6.00	14.80	2.70
13	6.50	15.20	2.63
14	7.00	15.50	1.94
15	7.50	15.80	1.90
16	8.00	16.10	1.86
17	8.50	16.40	1.83
18	9.00	16.70	1.80
19	9.50	17.00	1.76
20	10.00	17.30	1.73
21	10.50	17.50	1.14
22	11.00	17.90	2.23

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(17.9 - 30.90434) (1)/(1) = < 5.00 E-3 g/L Carbon

(17.9 - 30.90434) (1)/(1) (12) = < 4.17 E-4 Molar Carbon

<<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3957

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002967

Date: 03/12/99

Time: 11:13:41

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = .7 ug/minute C

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.50	66.67
3	1.51	7.00	78.57
4	2.00	47.80	85.36
5	2.50	126.40	62.18
6	3.00	216.00	41.48
7	3.50	329.00	34.35
8	4.00	449.90	26.87
9	4.50	516.40	12.88
10	5.00	544.80	5.21
11	5.50	555.30	1.89
12	6.00	559.60	0.77
13	6.50	561.90	0.41
14	7.00	563.20	0.23
15	7.50	564.60	0.25
16	8.00	565.40	0.14
17	8.50	566.30	0.16
18	9.00	567.40	0.19
19	9.50	568.10	0.12
20	10.00	568.70	0.11
21	10.50	569.60	0.16
22	11.00	570.10	0.09

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(570.1 - 7.698504) (1)/(1)	=	+5.624E+02	g/L Carbon
(570.1 - 7.698504) (1)/(1) (12)	=	+4.687E+01	Molar Carbon

Sample Run By:

SL HOOD

00002

SS = 100 uL = .1412g

3958

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002967

Date: 03/12/99

Time: 11:25:59

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.81 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==== Analysis Time ==== Coulometer ==== % Difference ==

1	0.51	0.60	0.00
2	1.01	1.30	53.85
3	1.51	10.30	87.38
4	2.00	39.90	74.19
5	2.50	109.20	63.46
6	3.00	187.40	41.73
7	3.50	236.00	20.59
8	4.00	261.10	9.61
9	4.50	273.20	4.43
10	5.00	280.10	2.46
11	5.50	284.00	1.37
12	6.00	286.90	1.01
13	6.50	288.90	0.69
14	7.00	290.50	0.55
15	7.50	291.80	0.45
16	8.00	292.80	0.34
17	8.50	293.60	0.27
18	9.00	294.70	0.37
19	9.50	296.10	0.47
20	10.00	297.30	0.40
21	10.50	298.00	0.23
22	11.00	298.70	0.23

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(298.7 - 30.89919) (1)/(1) = +2.678E+02 g/L Carbon
 (298.7 - 30.89919) (1)/(1) (12) = +2.232E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .100 ml = .1412g

3959

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002967 DW? Date: 03/12/99 Time: 11:42:59

Sample Size = 1 uL
Dil Factor = 1
Blank ID # =
Blank Value = .7 ug/minute C

Analyst : SL HOOD
Min Readings = 22
Max Readings = 22
% Difference = 10

== Reading ==	==== Analysis Time ==	==== Coulometer ==	==== % Difference ==
1	0.51	0.50	0.00
2	1.01	1.50	66.67
3	1.51	5.40	72.22
4	2.00	36.00	85.00
5	2.50	99.00	63.64
6	3.00	166.20	40.43
7	3.50	257.00	35.33
8	4.00	393.70	34.72
9	4.50	485.50	18.91
10	5.00	529.50	8.31
11	5.50	547.50	3.29
12	6.00	554.20	1.21
13	6.50	557.20	0.54
14	7.00	558.80	0.29
15	7.50	560.10	0.23
16	8.00	561.30	0.21
17	8.50	562.40	0.20
18	9.00	563.30	0.16
19	9.50	564.00	0.12
20	10.00	564.80	0.14
21	10.50	565.60	0.14
22	11.00	566.20	0.11

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(566.2 - 7.698504) (1)/(1) = +5.585E+02 g/L Carbon
(566.2 - 7.698504) (1)/(1) (12) = +4.654E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

35 = .100 ml = .1405g

3960

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002967 JUP Date: 03/12/99

Time: 11:56:27

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.81 ug/minute C

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference
1	0.51	0.60	0.00
2	1.01	1.50	60.00
3	1.51	18.70	91.98
4	2.01	61.90	69.79
5	2.50	138.30	55.24
6	3.00	209.90	34.11
7	3.50	252.10	16.74
8	4.00	272.90	7.62
9	4.50	283.00	3.57
10	5.00	288.20	1.80
11	5.50	291.80	1.23
12	6.00	294.00	0.75
13	6.50	295.70	0.57
14	7.00	296.80	0.37
15	7.50	298.00	0.40
16	8.00	299.00	0.33
17	8.50	299.80	0.27
18	9.00	300.50	0.23
19	9.50	301.20	0.23
20	10.00	301.80	0.20
21	10.50	302.40	0.20
22	11.00	302.90	0.17

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(302.9 - 30.90914) (1)/(1) = +2.720E+02 g/L Carbon
 (302.9 - 30.90914) (1)/(1) (12) = +2.267E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

3961

2B HALL

03/12/99 17:18 FAX 3721143

SS = .100 ml = .1405 g

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002968

Date: 03/12/99

Time: 13:39:56

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading	==== Analysis Time	==== Coulometer	==== % Difference ==
1	0.51	0.40	0.00
2	1.01	1.10	63.64
3	1.50	4.90	77.55
4	2.00	37.10	86.79
5	2.50	104.20	64.40
6	3.00	189.80	45.10
7	3.50	277.10	31.50
8	4.00	332.60	16.69
9	4.50	356.00	6.57
10	5.00	364.30	2.28
11	5.50	367.70	0.92
12	6.00	369.70	0.54
13	6.50	371.20	0.40
14	7.00	372.30	0.30
15	7.50	373.40	0.29
16	8.00	374.30	0.24
17	8.50	375.20	0.24
18	9.00	376.10	0.24
19	9.50	376.80	0.19
20	10.00	377.50	0.19
21	10.50	378.20	0.19
22	11.00	378.70	0.13

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 = +7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(378.7 - 7.69859) (1)/(1) = +3.710E+02 g/L Carbon
 (378.7 - 7.69859) (1)/(1) (12) = +3.092E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

SS = .100 ml = .146/g

3962

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T002968

Date: 03/12/99

Time: 13:54:10

Sample Size = 1 uL

Analyst : SL HOOD

Dil Factor = 1

Min Readings = 22

Blank ID # =

Max Readings = 22

Blank Value = 2.81 ug/minute C

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.70	0.00
2	1.01	1.80	61.11
3	1.51	27.40	93.43
4	2.00	100.30	72.68
5	2.50	178.80	43.90
6	3.00	230.60	22.46
7	3.50	257.80	10.55
8	4.00	270.20	4.59
9	4.50	276.40	2.24
10	5.00	279.90	1.25
11	5.50	282.60	0.96
12	6.00	284.50	0.67
13	6.50	285.90	0.49
14	7.00	286.90	0.35
15	7.50	288.00	0.38
16	8.00	288.90	0.31
17	8.50	289.60	0.24
18	9.00	290.50	0.31
19	9.50	291.10	0.21
20	10.00	291.70	0.21
21	10.50	292.50	0.27
22	11.00	293.20	0.24

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(293.2 - 30.89919) (1)/(1) = +2.623E+02 g/L Carbon
 (293.2 - 30.89919) (1)/(1) (12) = +2.186E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .100 ml = .1461g

3963

TIC- TOTAL INORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02968 DUP

Date: 03/12/99

Time: 14:19:45

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = .7 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.30	0.00
2	1.01	0.90	66.67
3	1.51	3.20	71.88
4	2.00	22.30	85.65
5	2.50	66.00	66.21
6	3.00	108.00	38.89
7	3.50	139.20	22.41
8	4.00	174.30	20.14
9	4.50	246.20	29.20
10	5.00	308.00	20.06
11	5.50	338.30	8.96
12	6.00	350.50	3.48
13	6.50	355.50	1.41
14	7.00	357.90	0.67
15	7.50	359.30	0.39
16	8.00	360.50	0.33
17	8.50	361.60	0.30
18	9.00	362.50	0.25
19	9.50	363.30	0.22
20	10.00	364.10	0.22
21	10.50	364.80	0.19
22	11.00	365.60	0.22

USER INPUT BLANK VALUE

BLANK VALUE = 7.698633 micrograms carbon

BLANK FACTOR = 7.698633 / 10.99805 =

+7.0E-01 ug/min Carbon

SAMPLE RESULTS:

(365.6 - 7.69859) (1)/(1) =

+3.579E+02 g/L Carbon

(365.6 - 7.69859) (1)/(1) (12) =

+2.983E+01 Molar Carbon

Sample Run By:

SL HOOD

00002

55 = .100 ml = .1423g

3964

TOC- TOTAL ORGANIC CARBON ANALYSIS REPORT
TICTOC REV 2.0

Sample: S98T02968 DUP

Date: 03/12/99

Time: 14:53:44

Sample Size = 1 uL

Dil Factor = 1

Blank ID # =

Blank Value = 2.81 ug/minute C

Analyst : SL HOOD

Min Readings = 22

Max Readings = 22

% Difference = 10

== Reading ==	== Analysis Time ==	== Coulometer ==	== % Difference ==
1	0.51	0.60	0.00
2	1.01	1.30	53.85
3	1.50	7.90	83.54
4	2.00	31.70	75.08
5	2.50	91.50	65.36
6	3.00	164.90	44.51
7	3.50	214.30	23.05
8	4.00	241.30	11.19
9	4.50	255.00	5.37
10	5.00	262.20	2.75
11	5.50	266.30	1.54
12	6.00	269.40	1.15
13	6.50	271.50	0.77
14	7.00	273.20	0.62
15	7.50	274.60	0.51
16	8.00	275.50	0.33
17	8.50	276.50	0.36
18	9.00	277.90	0.50
19	9.50	279.10	0.43
20	10.00	280.10	0.36
21	10.50	280.90	0.28
22	11.00	281.50	0.21

USER INPUT BLANK VALUE

BLANK VALUE = 30.90451 micrograms carbon

BLANK FACTOR = 30.90451 / 10.99805 = +2.8E+00 ug/min Carbon

SAMPLE RESULTS:

(281.5 - 30.91189) (1)/(1) = +2.506E+02 g/L Carbon
 (281.5 - 30.91189) (1)/(1) (12) = +2.088E+01 Molar Carbon
 <<<< WARNING - BLANK VALUE EXCEEDS 1.5 ug/min Carbon!!!!>>>>

Sample Run By:

SL HOOD

00002

SS = .100 uL = .1423g

3965

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 13 of 17

Document Information

Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 3966 THRU 4123)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	RADIOCHEMICAL ANALYSIS, TOTAL ALPHA ANALYSIS, AT WORKLIST		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSE Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2997)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

HNF-1663 REV. 0

RADIOCHEMICAL ANALYSIS

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

LABCORE Completed Worklist Report for Worklist# 26718

Analyst: scl

Instrument: AB15

Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: xBY-105 FOR @ALPHA01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 LIQUID		1.80E-04	1.67E-4	92.778 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E LIQUID		1.00	3.61E+00	3.610 % Ct. Error	
2 BLNK	0		@ALPHA01 ALPHA01 LIQUID		1	<7.58E-3		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01 LIQUID		1.00E+00	1.31E+00	1.310	BLNK/BKG
4 SAMPLE	S98T002806	0	@ALPHA01 ALPHA01 LIQUID		N/A	< 6.25E-3	9.89E-003	uCi/mL
4 SAMPLE	S98T002806	0	@ALPHA01 ALPHA01E LIQUID		N/A	5.00E+02		% Ct. Error
5 DUP	S98T002806	0	@ALPHA01 ALPHA01 LIQUID		<6.25E-3	<6.91E-3		RPD
5 DUP	S98T002806	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error
6 SAMPLE	S98T002807	0	@ALPHA01 ALPHA01 LIQUID		N/A	< 1.03E-2	1.98E-002	uCi/mL
6 SAMPLE	S98T002807	0	@ALPHA01 ALPHA01E LIQUID		N/A	5.00E+02		% Ct. Error
7 DUP	S98T002807	0	@ALPHA01 ALPHA01 LIQUID		<1.03E-2	<1.38E-2		RPD
7 DUP	S98T002807	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error
8 SAMPLE	S98T002808	0	@ALPHA01 ALPHA01 LIQUID		N/A	< 1.38E-2	1.98E-002	uCi/mL
8 SAMPLE	S98T002808	0	@ALPHA01 ALPHA01E LIQUID		N/A	2.34E+02		% Ct. Error
9 DUP	S98T002808	0	@ALPHA01 ALPHA01 LIQUID		<1.38E-2	<1.52E-2		RPD
9 DUP	S98T002808	0	@ALPHA01 ALPHA01E LIQUID		1.00	1.92E+02	192.000	% Ct. Error
10 SAMPLE	S98T002809	0	@ALPHA01 ALPHA01 LIQUID		N/A	1.77E-2	1.98E-002	uCi/mL
10 SAMPLE	S98T002809	0	@ALPHA01 ALPHA01E LIQUID		N/A	8.58E+01		% Ct. Error
11 DUP	S98T002809	0	@ALPHA01 ALPHA01 LIQUID		1.77E-2	1.77E-2	0.000	RPD
11 DUP	S98T002809	0	@ALPHA01 ALPHA01E LIQUID		1.00	7.67E+01	76.700	% Ct. Error
12 SAMPLE	S98T002810	0	@ALPHA01 ALPHA01 LIQUID		N/A	< 1.38E-2	1.98E-002	uCi/mL
12 SAMPLE	S98T002810	0	@ALPHA01 ALPHA01E LIQUID		N/A	2.34E+02		% Ct. Error
13 DUP	S98T002810	0	@ALPHA01 ALPHA01 LIQUID		<1.38E-2	<2.17E-2		RPD
13 DUP	S98T002810	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error

Final page for worklist# 26718

Analyst Signature

Date

3/7/98
Reviewer Signature Date

Analyst Signature

Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26718

Analyst: S.L Instrument: AB00 15 Book# 31857

Method: LA-508-101 Rev/Mod G-0

Worklist Comment: xBY-105 FOR @ALPHA01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@ALPHA01	LIQUID		
2	BLNK				@ALPHA01	LIQUID		
3	BLNK/BKG				@ALPHA01	LIQUID		
4	SAMPLE	S98T002806 0			@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
5	DUP	S98T002806 0			@ALPHA01	LIQUID		
6	SAMPLE	S98T002807 0			@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
7	DUP	S98T002807 0			@ALPHA01	LIQUID		
8	SAMPLE	S98T002808 0			@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
9	DUP	S98T002808 0			@ALPHA01	LIQUID		
10	SAMPLE	S98T002809 0			@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
11	DUP	S98T002809 0			@ALPHA01	LIQUID		
12	SAMPLE	S98T002810 0			@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
13	DUP	S98T002810 0			@ALPHA01	LIQUID		

Final page for worklist # 26718

Sue Lee 10-30-98
Analyst Signature Date

MB 11/1/98
Analyst Signature Date
Em Buda 11-2-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	15
STD	DISH SIZE (1, 2, or 5) (MS)	2
Worksheet	GROSS COUNTS (GC)	2970
26718	COUNT TIME in MINUTES (CT)	30
Alt or WB1	BACKGROUND in cpm (BKG)	0.33
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	98.670
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005324	Concentration in $\mu\text{Ci/L}$	1.65E-01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	1.69E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$	1.6687E-01
Sample/Rep		
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$	
Sample#	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
WL26718-STD	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Preparation		
MCB		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL} =$	1.67E-04
Analyst		
SCL		
Date/Complete	RELATIVE COUNTING ERROR =	3.6%
11/01/98		
Analysis Date		
10/30/98		
Analysis Time		
11:30 AM		
Sample Point		
BY-105		

Analyst:	MCB	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/1/98
STANDARD.WB1 Rev. 1.0	508101ML	

3970

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

LIQUIDS

		BLNK	REPLICATE
Type	DETECTOR NUMBER	15	15
BLNK	DISH SIZE (1 , 2 , or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	11	15
26718	COUNT TIME in MINUTES (CT)	30	30
ALPHA	BACKGROUND in cpm (BKG)	0.33	0.33
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.288	0.444

Batch Number	
98005324	Blank Concentration in $\mu\text{Ci/L}$ < 4.91E+00
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ < 7.58E+00
0	Maximum Concentration in $\mu\text{Ci/L}$ < 7.5780E+00

Sample Prep	N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	WL26718-BLNK	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 2220000dpm/ μCi)
Instrument Code	WB26872	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Prepared By	MCB	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
Chemist		Detection Levels and Less Than Values are determined from Procedure LA-508-002.

SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 7.58E-03	DETECTION LEVEL
Analyst	SCL		
Date Complete	11/01/98	RELATIVE COUNTING ERROR	500.0%

Analyst Date	10/30/98
Analyst Time	11:30 AM
Sample Point	BY-105

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:		SAC Date: 3/1/98
BLANK.WB1 Rev. 1.0	508101ML	

3971

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	13	10
26718	COUNT TIME in MINUTES (CT)	30	30
AT01157	BACKGROUND in cpm (BKG)	0.33	0.33
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
1037 Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.367	0.249
Batch Number			
98005324	Blank Concentration in $\mu\text{Ci/L}$	< 6.25E+00	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	< 4.24E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 6.2494E+00	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 222000\text{dpm}/\mu\text{Ci})$		
S98T002806	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 6.25E-03	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.		
SCL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	9.89E-03 $\mu\text{Ci/mL}$
11/01/98			
Analysis Date			
10/30/98			
Analysis Time			
11:30 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist: <i>John C. Cotton</i>	SAC	Date: <i>3 Nov 98</i>
SAMPLE.WB1 Rev. 1.0	508101ML	

3972

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0) LIQUIDS

		DUP	REPLICATE
TYPE	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Workfile	GROSS COUNTS (GC)	7	14
26718	COUNT TIME in MINUTES (CT)	30	30
At or 1157	BACKGROUND in cpm (BKG)	0.33	0.33
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.245	0.406
Batch Number			
98005324	Blank Concentration in $\mu\text{Ci/L}$	< 4.17E+00	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	< 6.91E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 6.9147E+00	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002806	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 6.91E-03	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
SCL			9.89E-03
Date Complete	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/mL}$
11/01/98			
Analysis Date			
10/30/98			
Analysis Time			
11:30 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/1/98

SAMPLE.WB1 Rev. 1.0

508101ML

3973

WORKBOOK PAGE: SAM6

AT: LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
WPC	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5)	2	2
Worklist	GROSS COUNTS	8	19
26718	COUNT TIME in MINUTES	30	30
At for 1572	BACKGROUND in cpm	0.33	0.33
AT	SAMPLE SIZE in mL	0.500	0.500
Test Code	DILUTION FACTOR	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR	1	1
Matrix	EFFICIENCY FACTOR	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.245	0.303

Batch Number	Blank Concentration in $\mu\text{Ci/L}$	< 8.35E+00
98005324	Replicate Concentration in $\mu\text{Ci/L}$	1.03E+01
Re-run	Maximum Concentration in $\mu\text{Ci/L}$	< 1.0344E+01
0		

Sample Prep	Rs (Sample Count Rate) = (TC / CT) - BKG
N/A	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 222000dpm/ μCi)
Sample #	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
S98T002807	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
Instrument Code	Detection Levels and Less Than Values are determined from Procedure LA-508-002.
WB26872	

Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.03E-02
Analyst	LESS Than Value was Determined from Lc.	
SCL		
Date Complete	RELATIVE COUNTING ERROR	500.0%
11/01/98		
Analysis Date		
10/30/98		
Analysis Time		
11:30 AM		
Sample Point		
BY-105		

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/26/98
SAMPLE.WB1 Rev. 1.0	508101ML	

3974

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

LIQUIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	11
26718	COUNT TIME in MINUTES (CT)	30
Ator 15?	BACKGROUND in cpm (BKG)	0.33
AT	SAMPLE SIZE in mL (SS)	0.500
1001 Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.288
Batch Number		
98005324	Blank Concentration in $\mu\text{Ci/L}$ <	9.82E+00
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	1.38E+01
0	Maximum Concentration in $\mu\text{Ci/L}$ <	1.3754E+01
Sample Prep		
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$	
Sample	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002807	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.38E-02
Analyst	LESS THAN Value was Determined from Rmax.	
SCL		
Date Complete	RELATIVE COUNTING ERROR	500.0%
11/01/98		
Analysis Date		
10/30/98		
Analysis Time		
11:30 AM		
Sample Point		
BY-105		

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/11/98
SAMPLE.WB1 Rev. 1.0 508101ML		

3975

WORKBOOK PAGE: SAM8

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5)	2	2
Worklist	GROSS COUNTS	14	14
26718	COUNT TIME in MINUTES	30	30
Avority?	BACKGROUND in cpm	0.33	0.33
AT	SAMPLE SIZE in mL	0.500	0.500
Inst Code	DILUTION FACTOR	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR	1	1
Matrix	EFFICIENCY FACTOR	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	0.406	0.406
BatchNumber			
98005324	Blank Concentration in $\mu\text{Ci/L}$	< 1.38E+01	
Retun	Replicate Concentration in $\mu\text{Ci/L}$	< 1.38E+01	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.3829E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample#	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 222000\text{dpm}/\mu\text{Ci})$		
S98T002808	ALPHA TOTAL $\mu\text{Ci/mL}$ = $ALPHA TOTAL \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.38E-02	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.		
SCL			1.98E-02
Date Complete	RELATIVE COUNTING ERROR	233.7%	$\mu\text{Ci/mL}$
11/01/98			
Analysis Date			
10/30/98			
Analysis Time			
11:30 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/1/98
SAMPLE.WB1 Rev. 1.0 508101ML		

3976

WORKBOOK PAGE: DUP9

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
WPC	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	15	22
26718	COUNT TIME in MINUTES (CT)	30	30
Alpha Br	BACKGROUND in cpm (BKG)	0.33	0.33
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.444	0.403

Batch Number	
98005324	Blank Concentration in $\mu\text{Ci/L}$ < 1.52E+01
Repur	Replicate Concentration in $\mu\text{Ci/L}$ 1.38E+01
0	Maximum Concentration in $\mu\text{Ci/L}$ < 1.5156E+01

Sample/Prep	
N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 2220000dpm/ μCi)
S98T002808	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
MCB	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) = < 1.52E-02
Analyst	
SCL	LESS THAN Value was Determined from Rmax.
Date/Complete	
11/01/98	RELATIVE COUNTING ERROR 191.8%
Analysis Date	
10/30/98	
Analysis Time	
11:30 AM	
Sample Point	
BY-105	

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	<i>SA Cath</i>	SAC Date: 3/2/98
SAMPLE.WB1 Rev. 1.0	508101ML	

3977

WORKBOOK PAGE: SAM10

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	23	28
26718	COUNT TIME in MINUTES (CT)	30	30
Al (O.I.B.)	BACKGROUND in cpm (BKG)	0.33	0.33
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Inst Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.437	0.603
Batch Number			
98005324	Blank Concentration in $\mu\text{Ci/L}$	1.49E+01	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	2.06E+01	
0	Average Concentration in $\mu\text{Ci/L}$	1.7732E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 222000\text{dpm}/\mu\text{Ci})$		
S98T002809	ALPHA TOTAL $\mu\text{Ci/mL}$ = $ALPHA\ TOTAL\ \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	1.77E-02	DETECTION LEVEL 1.98E-02 $\mu\text{Ci/mL}$
Analyst			
SCL			
Date/Complete	RELATIVE COUNTING ERROR	85.8%	
11/01/98			
Analysis Date			
10/30/98			
Analysis Time			
11:30 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/2/98

SAMPLE.WB1 Rev. 1.0

508101ML

3978

WORKBOOK PAGE: DUP11

AT : LA-508-101 (G-0)

LIQUIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	15
DUP	DISH SIZE (1, 2, or 5)	2
WorkList	GROSS COUNTS (MS)	25
26718	COUNT TIME in MINUTES (GC)	30
At or By	BACKGROUND in cpm (CT)	0.33
AT	SAMPLE SIZE in mL (BKG)	0.500
Test Code	DILUTION FACTOR (SS)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DF)	1
Matrix	EFFICIENCY FACTOR (DDF)	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE (EFF)	0.537
Batch Number	Blank Concentration in $\mu\text{Ci/L}$	1.83E+01
98005324	Replicate Concentration in $\mu\text{Ci/L}$	1.72E+01
Recon	Average Concentration in $\mu\text{Ci/L}$	1.7732E+01
0		
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002809	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	1.77E-02
Analyst		
SCL		
Date Complete		
11/01/98	RELATIVE COUNTING ERROR	76.7%
Analysis Date		
10/30/98		
Analysis Time		
11:30 AM		
Sample Point		
BY-105		

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/1/98
SAMPLE.WB1 Rev. 1.0	508T01ML	

3979

WORKBOOK PAGE: SAM12

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5)	2	2
WorkList	GROSS COUNTS	14	21
26718	COUNT TIME in MINUTES	30	30
At for 1517	BACKGROUND in cpm	0.33	0.33
AT	SAMPLE SIZE in mL	0.500	0.500
Test Code	DILUTION FACTOR	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR	1	1
Matrix	EFFICIENCY FACTOR	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.406	0.370

Batch Number	Blank Concentration in $\mu\text{Ci/L}$	< 1.38E+01
98005324	Replicate Concentration in $\mu\text{Ci/L}$	1.26E+01
ReRun	Maximum Concentration in $\mu\text{Ci/L}$	< 1.3829E+01
0		

Sample Prep	N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	S98T002810	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 222000dpm/ μCi)
Instrument Code	WB26872	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Prepared By	MCB	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
Chemist	SAC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum)	=	< 1.38E-02	DETECTION LEVEL
LESS THAN Value was Determined from Rmax.			
RELATIVE COUNTING ERROR		233.7%	1.98E-02 $\mu\text{Ci/mL}$

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/1/98
SAMPLE.WB1 Rev. 1.0	508101ML	

3980

WORKBOOK PAGE: DUP13

AT : LA-508-101 (G-0)

LIQUIDS

	DUP	REPLICATE
Type	15	15
DUP	2	2
Worklist	12	29
26718	30	30
At or After	0.33	0.33
AT	0.500	0.500
Test Code	10201	10201
@ALPHA01	1	1
Matrix	0.2695	0.2695
LIQUID	0.327	0.637
Batch Number		
98005324	< 1.12E+01	
ReRun	2.17E+01	
0	< 2.1711E+01	
Sample Prep		
N/A		
Sample		
S98T002810		
Instrument Code		
WB26872		
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 2.17E-02
Analyst		
SCL	LESS THAN Value was Determined from Rmax.	
Date Complete	RELATIVE COUNTING ERROR	500.0%
11/01/98		
Analysis Date		
10/30/98		
Analysis Time		
11:30 AM		
Sample Point		
BY-105		

Analyst:	SCL	Date: 01-Nov-98
Signature of Chemist:	SAC	Date: 3/11/98
SAMPLE.WB1 Rev. 1.0	508101ML	

3981

LABCORE Completed Worklist Report for Worklist# 26719

Analyst: crj

Instrument: AB17

Book#: _____

Method: LA-508-101 Rev/Mod _____

Worklist Comment: BY-105 FOR @ALPHA01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01	LIQUID	1.80E-04	1.76E-4	97.778 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E	LIQUID	1.00	3.30E+00	3.300 % Ct. Erro	
2 BLNK	0		@ALPHA01 ALPHA01	LIQUID	1	<1.26E-2		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01	LIQUID	1.00E+00	1.13E+00	1.130	BLNK/BKG
4 SAMPLE	S98T002811	0	@ALPHA01 ALPHA01	LIQUID	N/A <	1.03E-02	1.87E-002	uCi/mL
4 SAMPLE	S98T002811	0	@ALPHA01 ALPHA01E	LIQUID	N/A	5.00E+02		% Ct. Error
5 DUP	S98T002811	0	@ALPHA01 ALPHA01	LIQUID	<1.03E-2	<1.37E-2		RPD
5 DUP	S98T002811	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Erro
6 SAMPLE	S98T002812	0	@ALPHA01 ALPHA01	LIQUID	N/A <	1.49E-02	1.87E-002	uCi/mL
6 SAMPLE	S98T002812	0	@ALPHA01 ALPHA01E	LIQUID	N/A	5.00E+02		% Ct. Error
7 DUP	S98T002812	0	@ALPHA01 ALPHA01	LIQUID	<1.49E-2	<1.37E-2		RPD
7 DUP	S98T002812	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Erro
8 SAMPLE	S98T002813	0	@ALPHA01 ALPHA01	LIQUID	N/A <	1.14E-02	1.87E-002	uCi/mL
8 SAMPLE	S98T002813	0	@ALPHA01 ALPHA01E	LIQUID	N/A	5.00E+02		% Ct. Error
9 DUP	S98T002813	0	@ALPHA01 ALPHA01	LIQUID	<1.14E-2	<7.99E-3		RPD
9 DUP	S98T002813	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Erro
10 SAMPLE	S98T002938	0	@ALPHA01 ALPHA01	LIQUID	N/A <	1.49E-02	1.87E-002	uCi/mL
10 SAMPLE	S98T002938	0	@ALPHA01 ALPHA01E	LIQUID	N/A	2.11E+02		% Ct. Error
11 DUP	S98T002938	0	@ALPHA01 ALPHA01	LIQUID	<1.49E-2	<1.26E-2		RPD
11 DUP	S98T002938	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Erro
12 SPK	S98T002938	0	@ALPHA01 ALPHA01	LIQUID	4.05E-02	3.84E-02	94.815 % Recovery	
13 SAMPLE	S98T002955	0	@ALPHA01 ALPHA01	LIQUID	N/A <	4.00E-03	9.33E-003	uCi/mL
13 SAMPLE	S98T002955	0	@ALPHA01 ALPHA01E	LIQUID	N/A	5.00E+02		% Ct. Error
14 DUP	S98T002955	0	@ALPHA01 ALPHA01	LIQUID	<4.00E-3	<5.72E-3		RPD
14 DUP	S98T002955	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Erro
15 SPK	S98T002955	0	@ALPHA01 ALPHA01	LIQUID	4.05E-02	3.83E-02	94.568 % Recovery	

Final page for worklist# 26719

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26719

Analyst: CRJ Instrument: AB00 17 Book# 25857 ^{10/22/98} 31857

Method: LA-508-101 Rev/Mod C-D

Worklist Comment: BY-105 FOR @ALPHA01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@ALPHA01	LIQUID		
2	BLNK				@ALPHA01	LIQUID		
3	BLNK/BKG				@ALPHA01	LIQUID		
4	SAMPLE	S98T002811	0		@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested:			ALPHA01	, ALPHA01E		
5	DUP	S98T002811	0		@ALPHA01	LIQUID		
6	SAMPLE	S98T002812	0		@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested:			ALPHA01	, ALPHA01E		
7	DUP	S98T002812	0		@ALPHA01	LIQUID		
8	SAMPLE	S98T002813	0		@ALPHA01	LIQUID	98000437	BY-105 (2)
		Analytes Requested:			ALPHA01	, ALPHA01E		
9	DUP	S98T002813	0		@ALPHA01	LIQUID		
10	SAMPLE	S98T002938	0		@ALPHA01	LIQUID	98000490	BY-105 (2)
		Analytes Requested:			ALPHA01	, ALPHA01E		
11	DUP	S98T002938	0		@ALPHA01	LIQUID		
12	SPK	S98T002938	0		@ALPHA01	LIQUID		
13	SAMPLE	S98T002955	0		@ALPHA01	LIQUID	98000509	BY-105 (2)
		Analytes Requested:			ALPHA01	, ALPHA01E		
14	DUP	S98T002955	0		@ALPHA01	LIQUID		
15	SPK	S98T002955	0		@ALPHA01	LIQUID		

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

LABCORE Data Entry Template for Worklist# 26719

S Type	Sample#	R A	Test	Matrix	Group#	Project

Final page for worklist # 26719

Analyst Signature

Date

10/30/98

Analyst Signature

Date

11-2-98
10-2-98
11-3-98
11-2-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Detector type	DETECTOR NUMBER	17
STD	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	3559
26719	COUNT TIME in MINUTES (CT)	30
AT on TB ?	BACKGROUND in cpm (BKG)	0.4
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.3099
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	118.233
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005325	Concentration in $\mu\text{Ci/L}$ =	1.72E-01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.81E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.7641E-01
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
WL26719-STD	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.76E-04
Analyst		
CRJ		
Date Complete		
11/02/98	RELATIVE COUNTING ERROR =	3.3%
Analysis Date		
10/30/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

Analyst:	SLH2	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/26/98
STANDARD.WB1 Rev. 1.0	508101ML	

3985

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0) LIQUIDS

		BLNK	REPLICATE
Type	DETECTOR NUMBER	17	17
BLNK	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	11	16
26719	COUNT TIME IN MINUTES (CT)	30	30
Alfort B?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.269	0.424

Batch Number	98005325	Blank Concentration in $\mu\text{Ci/L}$	< 7.99E+00
Retun		Replicate Concentration in $\mu\text{Ci/L}$	< 1.26E+01
0		Maximum Concentration in $\mu\text{Ci/L}$	< 1.2585E+01

Sample Prep

N/A

Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$

Sample

ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$

WL26719-BLNK

ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$

Instrument Code

Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$

WB27804

Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By

SLH2

Chemist

SAC

ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) = < 1.26E-02

Analyst

CRJ

LESS Than Value was Determined from Lc.

Date Complete

11/02/98

RELATIVE COUNTING ERROR 500.0%

Analysis Date

10/30/98

Analysis Time

02:00 PM

Sample Point

BY-105 (2)

DETECTION
LEVEL1.87E-02
 $\mu\text{Ci/mL}$

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/24/98
BLANK.WB1 Rev. 1.0	508101ML	

3986

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	17	17
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	14	13
26719	COUNT TIME in MINUTES (CT)	30	30
A for B?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.347	0.308
Batch Number			
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.03E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 9.14E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.0294E+01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002811	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.03E-02	DETECTION LEVEL
Analyst			
CRJ	LESS THAN Value was Determined from Rmax.		
Date Complete	RELATIVE COUNTING ERROR	500.0%	1.87E-02 $\mu\text{Ci/mL}$
11/02/98			
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist: <i>SLH2</i>	SAC	Date: 3/2/99
SAMPLE.WB1 Rev. 1.	508101ML	

3987

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0) LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	17	17
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	17	13
26719	COUNT TIME in MINUTES (CT)	30	30
A or IB?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.463	0.308
Batch Number			
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.37E+01	
Return	Replicate Concentration in $\mu\text{Ci/L}$	< 9.14E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.3726E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002811	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.37E-02	DETECTION LEVEL
Analyst			
CRJ	LESS THAN Value was Determined from Rmax.		
Date Complete			1.87E-02
11/02/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/mL}$
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 30/11/98
SAMPLE.WB1 Rev. 1.	508101ML	

3988

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	17	17
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	18	11
26719	COUNT TIME in MINUTES (CT)	30	30
AT or B?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.501	0.269
Batch Number			
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.49E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 7.99E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.4865E+01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002812	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.49E-02	DETECTION LEVEL
Analyst			
CRJ	LESS THAN Value was Determined from Rmax.		
Date Complete			1.87E-02 $\mu\text{Ci/mL}$
11/02/98	RELATIVE COUNTING ERROR	500.0%	
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/2/98
SAMPLE.WB1 Rev. 1.	508101ML	

3989

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0) LIQUIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	17
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	17
26719	COUNT TIME in MINUTES (CT)	30
AT for TB?	BACKGROUND in cpm (BKG)	0.4
AT	SAMPLE SIZE in mL (SS)	0.500
Test Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.463
Batch Number		
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.37E+01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 7.99E+00
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.3726E+01
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002812	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[\{ (The Square Root of TC + BKG * CT) / (TC - BKG * CT) \}] * 1.96 * 100$	
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.37E-02
Analyst		
CRJ	LESS THAN Value was Determined from Rmax.	
Date Complete		
11/02/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date		
10/30/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

3990

WORKBOOK PAGE: SAM8

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	17	17
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	15	11
26719	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	0.386	0.269

Batch Number	98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.14E+01
Rerun	0	Replicate Concentration in $\mu\text{Ci/L}$	< 7.99E+00
		Maximum Concentration in $\mu\text{Ci/L}$	< 1.1441E+01

Sample Prep	N/A	$R_s \text{ (Sample Count Rate)} = (TC / CT) - BKG$	
Sample #	S98T002813	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
Instrument Code	WB27804	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$	
		$\text{Relative Counting Error} = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
Prepared By	SLH2	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Chemist	SAC		

SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) = < 1.14E-02			DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.			
CRJ				
Date Complete				1.87E-02
11/02/98	RELATIVE COUNTING ERROR 500.0%			$\mu\text{Ci/mL}$
Analysis Date				
10/30/98				
Analysis Time				
02:00 PM				
Sample Point				
BY-105 (2)				

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/16/98
SAMPLE.VB1 Rev. 1.	508101ML	

3991

WORKBOOK PAGE: DUP9

AT : LA-508-101 (G-0) LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	17	17
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	12	12
26719	COUNT TIME in MINUTES (CT)	30	30
Alon BY	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.269	0.269
Batch Number			
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 7.99E+00	
Repuns	Replicate Concentration in $\mu\text{Ci/L}$	< 7.99E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 7.9904E+00	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002813	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[ ](\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT) [ ] * 1.96 * 100$		
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 7.99E-03	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
CRJ			
Date Complete	RELATIVE COUNTING ERROR	500.0%	1.87E-02 $\mu\text{Ci/mL}$
11/02/98			
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/26/98
SAMPLE.WB1 Rev. 1. 508101ML		

3992

WORKBOOK PAGE: SAM10

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	17	17
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	18	17
26719	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs(SAMPLE RATE) as APPROPRIATE	0.501	0.463

Batch Number	98005325	Blank Concentration in $\mu\text{Ci/L}$	< 1.49E+01
Rerun		Replicate Concentration in $\mu\text{Ci/L}$	< 1.37E+01
0		Maximum Concentration in $\mu\text{Ci/L}$	< 1.4865E+01

Sample Prep	N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	S98T002938	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
Instrument Code	WB27804	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
		Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
		Detection Levels and Less Than Values are determined from Procedure LA-508-002.	

Prepared By	SLH2		
Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.49E-02
Analyst	CRJ	LESS THAN Value was Determined from Rmax.	DETECTION LEVEL
Date Complete	11/02/98	RELATIVE COUNTING ERROR	211.1%
Analysis Date	10/30/98		1.87E-02 $\mu\text{Ci/mL}$
Analysis Time	02:00 PM		
Sample Point	BY-105 (2)		

Analyst	CRJ	Date: 02-Nov-98
Signature of Chemist	SAC	Date: 3/26/98
SAMPLE.WB1 Rev. 1.	508101ML	

3993

WORKBOOK PAGE: DUP11

AT : LA-508-101 (G-0)   LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	17	17
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	16	8
26719	COUNT TIME in MINUTES (CT)	30	30
At Cont Bl?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.424	0.269

Batch Number	
98005325	Blank Concentration in $\mu\text{Ci/L}$ < 1.26E+01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 7.99E+00
0	Maximum Concentration in $\mu\text{Ci/L}$ < 1.2585E+01

Sample Prep	
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
S98T002938	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum)	=	< 1.26E-02
Analyst	LESS THAN Value was Determined from Rmax.		DETECTION LEVEL
CRJ			
Date Complete			1.87E-02
11/02/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/mL}$
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

3994

WORKBOOK PAGE: SPK12

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	17	17
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	78820	79548
26719	COUNT TIME in MINUTES (CT)	30	30
AT for IB?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.500	0.500
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
LIQUID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005325	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0546E-02	4.0546E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.3099	0.3099
0	SAMPLE + SPIKE $\mu\text{Ci/mL}$ (S+S)	7.79E+01	7.86E+01
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/mL}$ in SAMPLE	< 1.4865E-02	
N/A			
Sample #			
S98T002938			
Instrument Code	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
WB27804	SAMPLE + SPIKE $\mu\text{Ci/mL}$ = $Rs * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
Prepared By	QC ACTUAL = SVal		
SLH2	QC FOUND = $((S+S \mu\text{Ci/mL} - \text{SAMPLE } \mu\text{Ci/mL}) * ((SDF/SVol)/(DF*DDF/SS)))$		
Chemist	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
SAC			
Analyst			
CRJ			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/02/98			
Analysis Date			
10/30/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.84E-02	
02:00 PM	AVG. PERCENT SPIKE RECOVERY =	94.6%	
Sample Point			
BY-105 (2)			

Analyst:	SLH2	Date: 02-Nov-98
Signature of Chemist:	<i>SLH2</i>	SAC Date: 3/2/98
SPIKE.WB1 Rev. 1.0	508101ML	

3995

WORKBOOK PAGE: SAM13

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	17	17
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work Est	GROSS COUNTS (GC)	9	8
26719	COUNT TIME IN MINUTES (CT)	30	30
A for B?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.269	0.269
Batch Number			
98005325	Blank Concentration in $\mu\text{Ci/L}$	< 4.00E+00	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 4.00E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 3.9952E+00	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002955	$ALPHA \text{ TOTAL } \mu\text{Ci/mL} = ALPHA \text{ TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 4.00E-03	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
CRJ			9.33E-03
Date Complete	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/mL}$
11/02/98			
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

3996

WORKBOOK PAGE: DUP14

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	17	17
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	15	10
26719	COUNT TIME in MINUTES (CT)	30	30
At or B?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.3099	0.3099
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.386	0.269

Batch Number	
98005325	Blank Concentration in $\mu\text{Ci/L}$ < 5.72E+00
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 4.00E+00
0	Maximum Concentration in $\mu\text{Ci/L}$ < 5.7203E+00

Sample Prep	
N/A	$Rs \text{ (Sample Count Rate)} = (TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/L} = Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
S98T002955	$ALPHA \text{ TOTAL } \mu\text{Ci/mL} = ALPHA \text{ TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	$Relative \text{ Counting Error} = [(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27804	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SLH2	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) = < 5.72E-03
Analyst	
CRJ	LESS THAN Value was Determined from Rmax.
Date Complete	
11/02/98	RELATIVE COUNTING ERROR 500.0%
Analysis Date	
10/30/98	
Analysis Time	
02:00 PM	
Sample Point	
BY-105 (2)	

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

3997

WORKBOOK PAGE: SPK15

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	17	17
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	80687	77346
26719	COUNT TIME in MINUTES (CT)	30	30
At or TB?	BACKGROUND in cpm (BKG)	0.4	0.4
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	1.000	1.000
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
LIQUID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005325	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0546E-02	4.0546E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.3099	0.3099
0	SAMPLE + SPIKE $\mu\text{Ci/mL}$ (S+S)	3.99E+01	3.82E+01
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/mL}$ in SAMPLE	<	3.9952E-03
N/A			
Sample #			
S98T002955			
Instrument Code	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
WB27804	SAMPLE + SPIKE $\mu\text{Ci/mL}$ = $Rs * DF * DDF / (EFF * SS * 2220000 \text{dpm}/\mu\text{Ci})$		
Prepared By	QC ACTUAL = SVal		
SLH2	QC FOUND = $((S+S \mu\text{Ci/mL} - \text{SAMPLE } \mu\text{Ci/mL}) * ((SDF/SVol)/(DF*DDF/SS)))$		
Chemist	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
SAC			
Analyst			
CRJ			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/02/98			
Analysis Date			
10/30/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.83E-02	
02:00 PM	AVG. PERCENT SPIKE RECOVERY =	94.4%	
Sample Point			
BY-105 (2)			

Analyst:	SLH2	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 3/11/98
SPIKE.WB1 Rev. 1.0	508101ML	

3998

LABCORE Completed Worklist Report for Worklist# 26720

Analyst: crj

Instrument: AB14

Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: BY-105 FOR @ALPHA01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 LIQUID		1.80E-04	1.70E-4	94.444 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E LIQUID		1.00	3.59E+00	3.590 % Ct. Error	
2 BLNK	0		@ALPHA01 ALPHA01 LIQUID		1	<8.75E-3		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01 LIQUID		1.00E+00	1.24E+00	1.240	BLNK/BKG
4 SAMPLE	S98T002956	0	@ALPHA01 ALPHA01 LIQUID		N/A <	9.42E-03	1.11E-002	uCi/mL
4 SAMPLE	S98T002956	0	@ALPHA01 ALPHA01E LIQUID		N/A	3.40E+02		% Ct. Error
5 DUP	S98T002956	0	@ALPHA01 ALPHA01 LIQUID		<9.42E-3	<9.73E-3		RPD
5 DUP	S98T002956	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error
6 SAMPLE	S98T002957	0	@ALPHA01 ALPHA01 LIQUID		N/A	1.55E-02	2.22E-002	uCi/mL
6 SAMPLE	S98T002957	0	@ALPHA01 ALPHA01E LIQUID		N/A	1.27E+02		% Ct. Error
7 DUP	S98T002957	0	@ALPHA01 ALPHA01 LIQUID		1.55E-2	<1.26E-2		RPD
7 DUP	S98T002957	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error
8 SAMPLE	S98T002958	0	@ALPHA01 ALPHA01 LIQUID		N/A <	2.17E-02	2.22E-002	uCi/mL
8 SAMPLE	S98T002958	0	@ALPHA01 ALPHA01E LIQUID		N/A	2.14E+02		% Ct. Error
9 DUP	S98T002958	0	@ALPHA01 ALPHA01 LIQUID		<2.17E-2	1.43E-2		RPD
9 DUP	S98T002958	0	@ALPHA01 ALPHA01E LIQUID		1.00	1.07E+02	107.000	% Ct. Error
10 SAMPLE	S98T002959	0	@ALPHA01 ALPHA01 LIQUID		N/A	1.83E-02	2.22E-002	uCi/mL
10 SAMPLE	S98T002959	0	@ALPHA01 ALPHA01E LIQUID		N/A	8.78E+01		% Ct. Error
11 DUP	S98T002959	0	@ALPHA01 ALPHA01 LIQUID		1.83E-2	2.34E-2	24.460	RPD
11 DUP	S98T002959	0	@ALPHA01 ALPHA01E LIQUID		1.00	7.88E+01	78.800	% Ct. Error

Final page for worklist# 26720

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

W. C. Cuthbert
Reviewer Signature _____ Date 2 Nov 98

RPD of sample S98T002959 is well below the counting uncertainty. Results accepted.

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26720

Analyst:

CRJ

Instrument: AB00

14

Book#

28657 31857

Method: LA-508-101

Rev/Mod C1-D

Worklist Comment: BY-105 FOR @ALPHA01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@ALPHA01	LIQUID		
2	BLNK				@ALPHA01	LIQUID		
3	BLNK/BKG				@ALPHA01	LIQUID		
4	SAMPLE	S98T002956 0			@ALPHA01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
5	DUP	S98T002956 0			@ALPHA01	LIQUID		
6	SAMPLE	S98T002957 0			@ALPHA01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
7	DUP	S98T002957 0			@ALPHA01	LIQUID		
8	SAMPLE	S98T002958 0			@ALPHA01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
9	DUP	S98T002958 0			@ALPHA01	LIQUID		
10	SAMPLE	S98T002959 0			@ALPHA01	LIQUID	98000509	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
11	DUP	S98T002959 0			@ALPHA01	LIQUID		

Final page for worklist # 26720

[Signature] 10/30/98
Analyst Signature Date

[Signature] 11/2/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Detector Number	14	14
Dish Size (1, 2, or 5) (MS)	2	2
Gross Counts (GC)	3024	3106
Count Time in Minutes (CT)	30	30
Background in cpm (BKG)	0.43	0.43
Sample Size in mL (SS)	1.000	1.000
Dilution Factor (DF)	1	1
Standard Book Number (Std BN)	31B57	31B57
Efficiency Factor (EFF)	0.2692	0.2692
Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	100.370	103.103
Standard Value in $\mu\text{Ci/mL}$	1.80E-04	
Concentration in $\mu\text{Ci/L}$ =	1.68E-01	
Replicate Concentration in $\mu\text{Ci/L}$ =	1.73E-01	
Average Concentration in $\mu\text{Ci/L}$ =	1.7024E-01	

Sample Prep	N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	WL26720-STD	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Instrument Code	WB27807	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Prepared By	SLH2	$\text{Relative Counting Error} = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Chemist	SAC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.70E-04	DETECTION LEVEL
RELATIVE COUNTING ERROR =	3.6%	1.09E-06 $\mu\text{Ci/mL}$

Analyst:	SLH2	Date: 02-Nov-98
Signature of Chemist:		Date: 2/2/98
STANDARD.WB1 Rev. 1.0	508101ML	

4001

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

LIQUIDS

		BLNK	REPLICATE
Type	DETECTOR NUMBER	14	14
BLNK	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	23	9
26720	COUNT TIME in MINUTES (CT)	30	30
Attributed By	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.337	0.279

Batch Number	Blank Concentration in $\mu\text{Ci/L}$	5.75E+00
98005326	Replicate Concentration in $\mu\text{Ci/L}$	< 4.77E+00
Rerun	Maximum Concentration in $\mu\text{Ci/L}$	< 5.7467E+00
0		

Sample Prep	Rs (Sample Count Rate) = $(TC / CT) - BKG$
N/A	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Sample #	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.
Prepared By	
SLH2	

Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 5.75E-03	DETECTION LEVEL
Analyst	CRJ	LESS THAN Value was Determined from Rs.		1.11E-02
Date Complete	11/02/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/mL}$
Analysis Date	10/30/98			
Analysis Time	02:00 PM			
Sample Point	BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	<i>SLH2</i>	SAC Date: 2 Nov 98
BLANK.WB1 Rev. 1.0	508101ML	

4002

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

LIQUIDS

	SAMPLE	REPLICATE
Detector Type	DETECTOR NUMBER	14
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	20
26720	COUNT TIME in MINUTES (CT)	30
AT for (BY)	BACKGROUND in cpm (BKG)	0.43
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	0.552
Batch Number		
98005326	Blank Concentration in $\mu\text{Ci/L}$	< 9.42E+00
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 6.81E+00
0	Maximum Concentration in $\mu\text{Ci/L}$	< 9.4246E+00
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002956	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 9.42E-03
Analyst		
CRJ	LESS THAN Value was Determined from Rmax.	
Date Complete		
11/02/98	RELATIVE COUNTING ERROR	339.9%
Analysis Date		
10/30/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

DETECTION
LEVEL1.11E-02
 $\mu\text{Ci/mL}$

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 2 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4003

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

LIQUIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	14
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	10
26720	COUNT TIME in MINUTES (CT)	30
For IB	BACKGROUND in cpm (BKG)	0.43
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.279

Batch Number	98005326	Blank Concentration in $\mu\text{Ci/L}$	< 4.77E+00
Rerun	0	Replicate Concentration in $\mu\text{Ci/L}$	9.73E+00
Sample Prep		Maximum Concentration in $\mu\text{Ci/L}$	< 9.7295E+00

N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
S98T002956	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	SLH2	Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 9.73E-03	DETECTION LEVEL
Analyst	CRJ			LESS Than Value was Determined from Lc.		
Date Complete	11/02/98			RELATIVE COUNTING ERROR	500.0%	1.11E-02 $\mu\text{Ci/mL}$
Analysis Date	10/30/98					
Analysis Time	02:00 PM					
Sample Point	BY-105 (2)					

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	<i>SP Cotton</i>	SAC Date: 2 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4004

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	14	14
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	22	31
26720	COUNT TIME in MINUTES (CT)	30	30
A for B?	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.303	0.603
Batch Number			
98005326	Blank Concentration in $\mu\text{Ci/L}$	1.04E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	2.06E+01	
0	Average Concentration in $\mu\text{Ci/L}$	1.5476E+01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002957	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L		
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	1.55E-02	DETECTION LEVEL
Analyst			
CRJ			
Date Complete			2.22E-02 $\mu\text{Ci/mL}$
11/02/98	RELATIVE COUNTING ERROR	127.2%	
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 2/2/98
SAMPLE.WB1 Rev. 1.	508101ML	

4005

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	14	14
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	24	13
26720	COUNT TIME in MINUTES (CT)	30	30
Alpha Bkg	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.370	0.283
Batch Number			
98005326	Blank Concentration in $\mu\text{Ci/L}$	1.26E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 9.67E+00	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 1.2631E+01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002957	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.26E-02	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from R_s .		
CRJ			
Date Complete	RELATIVE COUNTING ERROR	500.0%	2.22E-02 $\mu\text{Ci/mL}$
11/02/98			
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 2/2/98
SAMPLE.WB1 Rev. 1.	508101ML	

4006

WORKBOOK PAGE: SAM8

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	14	14
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	32	18
26720	COUNT TIME in MINUTES (CT)	30	30
At For (B17)	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.637	0.476

Batch Number	
98005326	Blank Concentration in $\mu\text{Ci/L}$ 2.17E+01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 1.62E+01
0	Maximum Concentration in $\mu\text{Ci/L}$ < 2.1735E+01

Sample Prep	
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
S98T002958	$ALPHA \text{ TOTAL } \mu\text{Ci/mL} = ALPHA \text{ TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SLH2	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) = < 2.17E-02
Analysis	
CRJ	LESS THAN Value was Determined from R_s .
Date Complete	
11/02/98	RELATIVE COUNTING ERROR 213.6%
Analysis Date	
10/30/98	
Analysis Time	
02:00 PM	
Sample Point	
BY-105 (2)	

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	<i>St Cath</i>	SAC Date: 2/10/98
SAMPLE.WB1 Rev. 1.	508101ML	

4007

WORKBOOK PAGE: DUP9

AT : LA-508-101 (G-0) LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	14	14
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	27	24
26720	COUNT TIME in MINUTES (CT)	30	30
Atcor 1B1	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.470	0.370
Batch Number			
98005326	Blank Concentration in $\mu\text{Ci/L}$	1.60E+01	
Re-run	Replicate Concentration in $\mu\text{Ci/L}$	1.26E+01	
0	Average Concentration in $\mu\text{Ci/L}$	1.4338E+01	

Sample Prep	N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	S98T002958	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Instrument Code	WB27807	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Prepared By	SLH2	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Chemist	SAC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Analyst	CRJ	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	1.43E-02	DETECTION LEVEL
Date Complete	11/02/98	RELATIVE COUNTING ERROR	107.3%	2.22E-02 $\mu\text{Ci/mL}$
Analysis Date	10/30/98			
Analysis Time	02:00 PM			
Sample Point	BY-105 (2)			

Analyst	CRJ	Date: 02-Nov-98
Signature of Chemist	<i>[Signature]</i>	SAC Date: 27 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4008

WORKBOOK PAGE: SAM10

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	14	14
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	27	31
26720	COUNT TIME IN MINUTES (CT)	30	30
ALPHA B5	BACKGROUND IN cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.470	0.603

Batch Number	
98005326	Blank Concentration in $\mu\text{Ci/L}$ 1.60E+01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ 2.06E+01
0	Average Concentration in $\mu\text{Ci/L}$ 1.8321E+01

Sample Prep	
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
S98T002959	$ALPHA \text{ TOTAL } \mu\text{Ci/mL} = ALPHA \text{ TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SLH2	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) = 1.83E-02
Analyst	
CRJ	
Date Complete	
11/02/98	RELATIVE COUNTING ERROR 87.8%
Analysis Date	
10/30/98	
Analysis Time	
02:00 PM	
Sample Point	
BY-105 (2)	

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:		SAC Date: 2/11/98
SAMPLE.WB1 Rev. 1.	508101ML	

4009

WORKBOOK PAGE: DUP11

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	14	14
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	29	38
26720	COUNT TIME in MINUTES (CT)	30	30
AT for TB	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2692	0.2692
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.537	0.837
Batch Number			
98005326	Blank Concentration in $\mu\text{Ci/L}$	1.83E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	2.86E+01	
0	Average Concentration in $\mu\text{Ci/L}$	2.3442E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002959	ALPHA TOTAL $\mu\text{Ci/mL}$ = $ALPHA TOTAL \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27807	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	2.34E-02	DETECTION LEVEL
Analyst			
CRJ			
Date Complete			2.22E-02
11/02/98	RELATIVE COUNTING ERROR	78.8%	$\mu\text{Ci/mL}$
Analysis Date			
10/30/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 02-Nov-98
Signature of Chemist:	SAC	Date: 2 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4010

LABCORE Completed Worklist Report for Worklist# 26838

Analyst: scl Instrument: AB18 Book#

Method: Rev/Mod

Worklist Comment: BY-105 FOR @ALPHA01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 LIQUID		1.80E-04	1.69E-4	93.889 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E LIQUID		1.00	3.62E+00	3.620 % Ct. Erro	
2 BLNK	0		@ALPHA01 ALPHA01 LIQUID		1	<9.21E-3		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E LIQUID		1.00	1.96E+02	196.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01 LIQUID		1.00E+00	2.00E+00	2.000	BLNK/BKG
4 SAMPLE	S98T003002	0	@ALPHA01 ALPHA01 LIQUID		N/A	4.92E-02	1.25e-002	uCi/mL
4 SAMPLE	S98T003002	0	@ALPHA01 ALPHA01E LIQUID		N/A	3.59E+01		% Ct. Error
5 DUP	S98T003002	0	@ALPHA01 ALPHA01 LIQUID		4.92E-2	4.46E-2		9.808 RPD
5 DUP	S98T003002	0	@ALPHA01 ALPHA01E LIQUID		1.00	3.92E+01	39.200	% Ct. Erro
6 SAMPLE	S98T003003	0	@ALPHA01 ALPHA01 LIQUID		N/A	4.58E-02	1.25e-002	uCi/mL
6 SAMPLE	S98T003003	0	@ALPHA01 ALPHA01E LIQUID		N/A	3.42E+01		% Ct. Error
7 DUP	S98T003003	0	@ALPHA01 ALPHA01 LIQUID		4.58E-2	4.11E-2		10.817 RPD
7 DUP	S98T003003	0	@ALPHA01 ALPHA01E LIQUID		1.00	3.71E+01	37.100	% Ct. Erro
8 SAMPLE	S98T003004	0	@ALPHA01 ALPHA01 LIQUID		N/A	4.75E-02	1.25e-002	uCi/mL
8 SAMPLE	S98T003004	0	@ALPHA01 ALPHA01E LIQUID		N/A	3.47E+01		% Ct. Error
9 DUP	S98T003004	0	@ALPHA01 ALPHA01 LIQUID		4.75E-2	4.07E-2		2.495 RPD
9 DUP	S98T003004	0	@ALPHA01 ALPHA01E LIQUID		1.00	3.28E+01	32.800	% Ct. Erro
10 SPK	S98T003004	0	@ALPHA01 ALPHA01 LIQUID		4.05E-02	3.48E-02		85.926 % Recovery
11 SAMPLE	S98T003005	0	@ALPHA01 ALPHA01 LIQUID		N/A	4.92E-02	1.25e-002	uCi/mL
11 SAMPLE	S98T003005	0	@ALPHA01 ALPHA01E LIQUID		N/A	3.53E+01		% Ct. Error
12 DUP	S98T003005	0	@ALPHA01 ALPHA01 LIQUID		4.92E-2	6.02E-2		20.110 RPD
12 DUP	S98T003005	0	@ALPHA01 ALPHA01E LIQUID		1.00	2.93E+01	29.300	% Ct. Erro

Final page for worklist# 26838

Analyst Signature Date

Analyst Signature Date

Reviewer Signature Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26838

Analyst: S.L Instrument: AB00 #18 Book# 31B57

Method: LA-508-101 Rev/Mod GT-0

Worklist Comment: BY-105 FOR @ALPHA01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@ALPHA01	LIQUID		
2	BLNK				@ALPHA01	LIQUID		
3	BLNK/BKG				@ALPHA01	LIQUID		
4	SAMPLE	S98T003002 0			@ALPHA01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
5	DUP	S98T003002 0			@ALPHA01	LIQUID		
6	SAMPLE	S98T003003 0			@ALPHA01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
7	DUP	S98T003003 0			@ALPHA01	LIQUID		
8	SAMPLE	S98T003004 0			@ALPHA01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
9	DUP	S98T003004 0			@ALPHA01	LIQUID		
10	SPK	S98T003004 0			@ALPHA01	LIQUID		
11	SAMPLE	S98T003005 0			@ALPHA01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
12	DUP	S98T003005 0			@ALPHA01	LIQUID		

Final page for worklist # 26838

Sue Lan 10-29-98
Analyst Signature Date

[Signature] 10-30-98
Analyst Signature Date
[Signature] 10/31/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	18
STD	DISH SIZE (1, 2, or 5) (MS)	2
Worklist	GROSS COUNTS (GC)	2946
26838	COUNT TIME in MINUTES (CT)	30
Alt or life?	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	98.100
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005442	Concentration in $\mu\text{Ci/L}$ =	1.67E-01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.70E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.6875E-01
Sample Prep		
N/A		
Sample #	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
WL26838-STD	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
Instrument Code	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
WB27809	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
Prepared By	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
VAR		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.69E-04
Analyst		
SCL		
Date Complete		
10/30/98	RELATIVE COUNTING ERROR =	3.6%
Analysis Date		
10/29/98		
Analysis Time		
03:20 PM		
Sample Point		
BY-105 (2)		

Analyst:	VAR	Date: 30-Oct-98
Signature of Chemist:	SAC	Date: 27/1/98
STANDARD.WB1 Rev. 1.0 508101ML		

4013

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

LIQUIDS

		BLNK	REPLICATE
WPP	DETECTOR NUMBER	18	18
BLNK	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkBt	GROSS COUNTS (GC)	6	6
26838	COUNT TIME in MINUTES (CT)	30	30
Av or Bt	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.265	0.265

Batch Number	
98005442	Blank Concentration in $\mu\text{Ci/L}$ < 9.21E+00
Roxin	Replicate Concentration in $\mu\text{Ci/L}$ < 9.21E+00
0	Maximum Concentration in $\mu\text{Ci/L}$ < 9.2109E+00

Sample Prep	
N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 2220000dpm/ μCi)
WL26838-BLNK	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum)	=	< 9.21E-03
Analyst	LESS THAN Value was Determined from Rmax.		DETECTION
SCL			LEVEL
Date Complete	RELATIVE COUNTING ERROR	196.0%	1.25E-02
10/30/98			$\mu\text{Ci/mL}$
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist:	SAC	Date: 27/10/98
BLANK.WB1 Rev. 1.0	508101ML	

4014

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worksheet	GROSS COUNTS (GC)	38	53
26838	COUNT TIME in MINUTES (CT)	30	30
At or By	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
1051000	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	1.167	1.667

Batch Number	
98005442	Blank Concentration in $\mu\text{Ci/L}$ 4.06E+01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ 5.79E+01
0	Average Concentration in $\mu\text{Ci/L}$ 4.9241E+01

Sample Prep	N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	S98T003002	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 2220000dpm/ μCi)
Instrument Code	WB27809	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
		Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
		Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	VAR		
Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.92E-02
Analyst	SCL		DETECTION LEVEL
Date Complete	10/30/98	RELATIVE COUNTING ERROR	35.9%
Analysis Date	10/29/98		1.25E-02 $\mu\text{Ci/mL}$
Analysis Time	03:20 PM		
Sample Point	BY-105 (2)		

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist:	SAC	Date: 27/10/98
SAMPLE.WB1 Rev. 1.0	508101ML	

4015

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	18	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work Unit	GROSS COUNTS (GC)	33	50
26838	COUNT TIME in MINUTES (CT)	30	30
Amount	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.000	1.567

Batch Number	
98005442	Blank Concentration in $\mu\text{Ci/L}$ 3.48E+01
Run	Replicate Concentration in $\mu\text{Ci/L}$ 5.45E+01
0	Average Concentration in $\mu\text{Ci/L}$ 4.4606E+01

Sample Prep	Rs (Sample Count Rate) = $(TC / CT) - BKG$
N/A	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Sample #	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
S98T003002	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Instrument Code	Detection Levels and Less Than Values are determined from Procedure LA-508-002.
WB27809	

Prepared By	VAR	
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) = 4.46E-02	DETECTION LEVEL
Analyst		
SCL		
Date Complete	RELATIVE COUNTING ERROR 39.2%	1.25E-02
10/30/98		$\mu\text{Ci/mL}$

Analysis Date	10/29/98
Analysis Time	03:20 PM
Sample Point	BY-105 (2)

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist:	<i>S. A. Nathan</i>	SAC Date: 27 Oct 98
SAMPLE.WB1 Rev. 1.0	508101ML	

4016

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	44	41
26838	COUNT TIME in MINUTES (CT)	30	30
Alpha B?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	1.367	1.267
Batch Number			
98005442	Blank Concentration in $\mu\text{Ci/L}$	4.76E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	4.40E+01	
0	Average Concentration in $\mu\text{Ci/L}$	4.5765E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003003	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared by			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.58E-02	DETECTION LEVEL
Analyst			
SCL			
Date Complete			1.25E-02
10/30/98	RELATIVE COUNTING ERROR	34.2%	$\mu\text{Ci/mL}$
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist: <i>SA Coath</i>	SAC	Date: <i>27/10/98</i>
SAMPLE.WB1 Rev. 1.0	508101ML	

4017

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

LIQUIDS

	DUP	REPLICATE
TYPE	DETECTOR NUMBER	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2
WorkSheet	GROSS COUNTS (GC)	41
26838	COUNT TIME in MINUTES (CT)	30
AT or ET	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	0.500
Test Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.267
Batch Number		1.100
98005442	Blank Concentration in $\mu\text{Ci/L}$	4.40E+01
Re-run	Replicate Concentration in $\mu\text{Ci/L}$	3.82E+01
0	Average Concentration in $\mu\text{Ci/L}$	4.1131E+01
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
S98T003003	ALPHA TOTAL $\mu\text{Ci/mL}$ = $ALPHA\ TOTAL\ \mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
VAR		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.11E-02
Analyst		
SCL		
Date/Complete	RELATIVE COUNTING ERROR	37.1%
10/30/98		
Analysis Date		
10/29/98		
Analysis Time		
03:20 PM		
Sample Point		
BY-105 (2)		

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist: <i>A. A. Cath</i>	SAC	Date: <i>2/10/98</i>
SAMPLE.WB1 Rev. 1.0	508101ML	

4018

WORKBOOK PAGE: SAM8

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkSheet	GROSS COUNTS (GC)	48	40
26838	COUNT TIME in MINUTES (CT)	30	30
Arion Bk	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.500	1.233
Batch Number			
98005442	Blank Concentration in $\mu\text{Ci/L}$	5.21E+01	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	4.29E+01	
0	Average Concentration in $\mu\text{Ci/L}$	4.7503E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003004	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.75E-02	DETECTION LEVEL
Analyst			
SCL			
Date Complete			1.25E-02 $\mu\text{Ci/mL}$
10/30/98	RELATIVE COUNTING ERROR	34.7%	
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist: <i>[Signature]</i>	SAC	Date: 27/10/98

SAMPLE.WB1 Rev. 1.0

508101ML

4019

WORKBOOK PAGE: DUP9

AT : LA-508-101 (G-0) LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	18	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	44	46
26838	COUNT TIME in MINUTES (CT)	30	30
AT for BY	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	1.367	1.433
Batch/Number			
98005442	Blank Concentration in $\mu\text{Ci/L}$	4.75E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	4.98E+01	
0	Average Concentration in $\mu\text{Ci/L}$	4.8662E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003004	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.87E-02	DETECTION LEVEL 1.25E-02 $\mu\text{Ci/mL}$
Analyst			
SCL			
Date Complete	RELATIVE COUNTING ERROR	32.8%	
10/30/98			
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist: <i>SAC</i>	SAC	Date: <i>27 Oct 98</i>
SAMPLE.WB1 Rev. 1.0 508101ML		

4020

WORKBOOK PAGE: SPK10

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
TYPE	DETECTOR NUMBER	18	18
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
WorkList	TOTAL COUNTS (TC)	61334	61212
26838	COUNT TIME in MINUTES (CT)	30	30
WorkList	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.500	0.500
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
LIQUID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005442	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0546E-02	4.0546E-02
Run	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2644	0.2644
0	SAMPLE + SPIKE $\mu\text{Ci/mL}$ (S+S)	7.11E+01	7.09E+01
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/mL}$ in SAMPLE	4.7503E-02	
N/A			
Sample #			
S98T003004			
Instrument Code	Rs (Sample Count Rate) = (TC / CT) - BKG		
WB27809	SAMPLE + SPIKE $\mu\text{Ci/mL}$ = Rs * DF * DDF / (EFF * SS * 2220000dpm/ μCi)		
Prepared By	QC ACTUAL = SVal		
VAR	QC FOUND = (((S+S $\mu\text{Ci/mL}$ - SAMPLE $\mu\text{Ci/mL}$) * ((SDF/SVol)/(DF*DDF/SS)))		
Chemist	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
SAC			
Analyst			
SCL			
Date Complete			
10/30/98			
Analysis Date			
10/29/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.48E-02	
03:20 PM	AVG. PERCENT SPIKE RECOVERY =	85.8%	
Sample Point			
BY-105 (2)			

Analyst:	VAR	Date: 30-Oct-98
Signature of Chemist:	SAC	Date: 2/1/98
SPIKE.WB1 Rev. 1.0 508101ML		

4021

WORKBOOK PAGE: SAM11

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	39	52
26838	COUNT TIME in MINUTES (CT)	30	30
Attn: Jeff	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	1.200	1.633
Batch Number			
98005442	Blank Concentration in $\mu\text{Ci/L}$	4.17E+01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	5.68E+01	
0	Average Concentration in $\mu\text{Ci/L}$	4.9241E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = (TC / CT) - BKG		
Sample /	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003005	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L		
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	4.92E-02	DETECTION LEVEL 1.25E-02 $\mu\text{Ci/mL}$
Analyst			
SCL			
Date Complete	RELATIVE COUNTING ERROR	35.3%	
10/30/98			
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist: <i>10.056</i>	SAC	Date: <i>2/1/98</i>
SAMPLE.WB1 Rev. 1.0	508101ML	

4022

WORKBOOK PAGE: DUP12

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	18	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	53	57
26838	COUNT TIME in MINUTES (CT)	30	30
At of By	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	1.667	1.800
Batch Number			
98005442	Blank Concentration in $\mu\text{Ci/L}$	5.79E+01	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	6.26E+01	
0	Average Concentration in $\mu\text{Ci/L}$	6.0248E+01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003005	ALPHA TOTAL $\mu\text{Ci/mL}$ = $\text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
VAR			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Average) =	6.02E-02	DETECTION LEVEL
Analyst			
SCL			
Date Complete	RELATIVE COUNTING ERROR	29.3%	1.25E-02 $\mu\text{Ci/mL}$
10/30/98			
Analysis Date			
10/29/98			
Analysis Time			
03:20 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 30-Oct-98
Signature of Chemist:	SAC	Date: 27/10/98
SAMPLE.WB1 Rev. 1.0 508101ML		

4023

LABCORE Completed Worklist Report for Worklist# 27092

Analyst: scl Instrument: AB13 Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: Rerun of WL 26965,BY-105 mcb

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@ALPHA01 ALPHA01	SOLID	1.80E-04	1.67E-4	92.778 % Recovery	
1 STD	0	@ALPHA01 ALPHA01E	SOLID	1.00	3.42E+00	3.420 % Ct. Error	
2 BLNK-PREP	0	@ALPHA01 ALPHA01	SOLID	1	<2.03E-1	uCi/g	
2 BLNK-PREP	0	@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000 uCi/g	
3 BLNK/BKG	0	@ALPHA01 ALPHA01	SOLID	1.00E+00	1.45E+00	1.450 BLNK/BKG	
4 SAMPLE	S98T002598	0 F @ALPHA01 ALPHA01	SOLID	N/A	1.54E-01	0.375 uCi/g	
4 SAMPLE	S98T002598	0 F @ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02	% Ct. Error	
5 DUP	S98T002598	0 F @ALPHA01 ALPHA01	SOLID	<1.54E-1	<1.87E-1	RPD	
5 DUP	S98T002598	0 F @ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000 % Ct. Error	
6 SPK	S98T002598	0 F @ALPHA01 ALPHA01	SOLID	4.05E-02	3.40E-02	83.951 % Recovery	
7 SAMPLE	S98T002599	0 F @ALPHA01 ALPHA01	SOLID	N/A	2.45E-01	0.371 uCi/g	
7 SAMPLE	S98T002599	0 F @ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02	% Ct. Error	
8 DUP	S98T002599	0 F @ALPHA01 ALPHA01	SOLID	<2.45E-1	<1.05E-1	RPD	
8 DUP	S98T002599	0 F @ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000 % Ct. Error	
9 SAMPLE	S98T002600	0 F @ALPHA01 ALPHA01	SOLID	N/A	2.05E-01	0.379 uCi/g	
9 SAMPLE	S98T002600	0 F @ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02	% Ct. Error	
10 DUP	S98T002600	0 F @ALPHA01 ALPHA01	SOLID	<2.05E-1	<1.59E-1	RPD	
10 DUP	S98T002600	0 F @ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000 % Ct. Error	

Final page for worklist# 27092

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature *John Reilly* Date *12/16/98*

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27092

Analyst: S.L Instrument: AB00 13 Book# 21B57

Method: LA-508-101 Rev/Mod G7-0

Worklist Comment: Rerun of WL 26965,BY-105 mcb

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T002598	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
5 DUP	S98T002598	0 F	@ALPHA01	SOLID		
6 SPK	S98T002598	0 F	@ALPHA01	SOLID		
7 SAMPLE	S98T002599	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
8 DUP	S98T002599	0 F	@ALPHA01	SOLID		
9 SAMPLE	S98T002600	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
10 DUP	S98T002600	0 F	@ALPHA01	SOLID		

Final page for worklist # 27092

Sue Lai 11-10-98
Analyst Signature Date

M.C. Brown 11/14/98
Analyst Signature Date
Sylvia Z. Chuan 11/11/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
TYPE	DETECTOR NUMBER	13
STD	DISH SIZE (1, 2, or 5) (MS)	2
Work Book	GROSS COUNTS (GC)	3298
27092	COUNT TIME in MINUTES (CT)	30
At of HB ?	BACKGROUND in cpm (BKG)	0.23
AT	SAMPLE SIZE in mL (SS)	1.000
test code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2971
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	109.703
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005701	Concentration in $\mu\text{Ci/L}$ =	1.66E-01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.67E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.6676E-01

Sample Prep N/A
 Sample # WL27092-STD
 Instrument Code WB27810
 Prepared By MCB
 Chemist
 SAC
 Analyst
 SCL
 Date Complete 11/11/98
 Analysis Date 11/10/98
 Analysis Time 09:45 PM
 Sample Point BY-105(2)

Rs (Sample Count Rate) = $(TC / CT) - BKG$
 $\text{ALPHA TOTAL } \mu\text{Ci/L} = \text{Rs} * 1000\text{mL/L} * DF / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$
 $\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
 $\text{Relative Counting Error} = [|(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)|] * 1.96 * 100$
 Detection Levels and Less Than Values are determined from Procedure LA-508-002.

ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.67E-04	DETECTION LEVEL 7.57E-07 $\mu\text{Ci/mL}$
RELATIVE COUNTING ERROR =	3.4%	

Analyst:	MCB	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98

STANDARD.WB1 Rev. 1.0 508101ML

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

		BLNK-PREP	REPLICATE
1709	DETECTOR NUMBER	13	13
BLNK-PREP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work Unit	GROSS COUNTS (GC)	15	5
27092	COUNT TIME in MINUTES (CT)	30	30
AT 0711/98	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0372	2.0372
17092	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.270	0.204

Blank Number		
98005701	Blank Concentration in $\mu\text{Ci/g}$	2.03E-01
Rep. In	Replicate Concentration in $\mu\text{Ci/g}$	< 1.54E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.0295E-01

Sample Prep	
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$

WL27092-BLNK	
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By				
MCB				
Chemist				
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 2.03E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rs.			
SCL				
Date Complete	RELATIVE COUNTING ERROR		500.0%	3.75E-01 $\mu\text{Ci/g}$
11/11/98				

Analysis Date	
11/10/98	
Analysis Time	
09:45 PM	
Sample Point	
BY-105(2)	

Analyst	SCL	Date: 11-Nov-98
Signature of Chemist: <i>John R. Kelly</i>	SAC	Date: 12 Nov 98
BLANK.WB1 Rev. 1.0		

4027

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
WFO	DETECTOR NUMBER	13	13
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WFO	GROSS COUNTS (GC)	6	5
27092	COUNT TIME in MINUTES (CT)	30	30
AT or TB	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0372	2.0372
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.204	0.204

Batch Number		
98005701	Blank Concentration in $\mu\text{Ci/g}$	< 1.54E-01
Recon	Replicate Concentration in $\mu\text{Ci/g}$	< 1.54E-01
1	Maximum Concentration in $\mu\text{Ci/g}$	< 1.5358E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample Of	$ALPHA\ TOTAL\ \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002598	
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 1.54E-01
Analyst	LESS Than Value was Determined from Lc.		DETECTION LEVEL
SCL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	3.75E-01 $\mu\text{Ci/g}$
11/11/98			
Analysis Date			
11/10/98			
Analysis Time			
09:45 PM			
Sample Point			
BY-105(2)			

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	SAC Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0	608101ML	

4028

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
DETECTOR NUMBER	13	13
DUP	2	2
DISH SIZE (1, 2, or 5) (MS)	2	2
GROSS COUNTS (GC)	2	8
COUNT TIME in MINUTES (CT)	30	30
BACKGROUND in cpm (BKG)	0.23	0.23
SAMPLE SIZE in mL (SS)	0.100	0.100
DILUTION FACTOR (DF)	101	101
DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0364	2.0364
EFFICIENCY FACTOR (EFF)	0.2971	0.2971
Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.204	0.249
Blank Concentration in $\mu\text{Ci/g}$	< 1.54E-01	
Replicate Concentration in $\mu\text{Ci/g}$	< 1.87E-01	
Maximum Concentration in $\mu\text{Ci/g}$	< 1.8722E-01	
Rs (Sample Count Rate) = (TC / CT) - BKG		
ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)		
Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100		
Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.87E-01	DETECTION LEVEL 3.75E-01 $\mu\text{Ci/g}$
LESS Than Value was Determined from Lc.		
RELATIVE COUNTING ERROR	500.0%	

11/11/98
 11/10/98
 09:45 PM
 BY-105(2)

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0		

WORKBOOK PAGE: SPK6

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
WFO	DETECTOR NUMBER	13	13
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Worklist	TOTAL COUNTS (TC)	68431	65996
27092	COUNT TIME in MINUTES (CT)	30	30
27092	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
27092	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0372	2.0372
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005701	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Recon	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2971	0.2971
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.71E+03	1.65E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	< 1.5358E-01	
FUSION01			
Sample #			
S98T002598	Rs (Sample Count Rate) = (TC / CT) - BKG		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / ((EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci}))$		
WB27810	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000)))/(DF/SS/Dg/L)))$		
MCB	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
Chemist			
SAC			
Analyst			
SCL			
Data Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/11/98			
Analysis Date			
11/10/98	QC ACTUAL =	4.05E-02	
Analysis time	QC FOUND =	3.40E-02	
09:45 PM	AVG. PERCENT SPIKE RECOVERY =	83.8%	
Sample Point			
BY-105(2)			

Analyst:	MCB	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	SAC Date: 12-Nov-98
SPIKE.WB1 Rev. 1.0	508101ML	

4030

WORKBOOK PAGE: SAM7

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Wt%	DETECTOR NUMBER	13	13
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Wt% (G)	GROSS COUNTS (GC)	6	10
27092	COUNT TIME in MINUTES (CT)	30	30
At or By	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0608	2.0608
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	0.204	0.329

Batch Number	
98005701	Blank Concentration in $\mu\text{Ci/g}$ < 1.52E-01
Run	Replicate Concentration in $\mu\text{Ci/g}$ < 2.45E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 2.4479E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002599	

Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
MCB	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 2.45E-01
Analyst	
SCL	LESS Than Value was Determined from Lc.
Date Complete	
11/11/98	RELATIVE COUNTING ERROR 500.0%
Analysis Date	
11/10/98	
Analysis Time	
09:45 PM	
Sample Point	
BY-105(2)	

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Regea</i>	SAC
SAMPLE.WB1 Rev. 1.0	508101ML	Date: 12 Nov 98

4031

WORKBOOK PAGE: DUP8

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
WPL	DETECTOR NUMBER	13	13
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	12	2
27092	COUNT TIME in MINUTES (CT)	30	30
Autobk	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0564	2.0564
Math	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.409	0.204

Batch Number	
98005701	Blank Concentration in $\mu\text{Ci/g}$ < 3.05E-01
ReRun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.52E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 3.0465E-01

Sample Prep	
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002599	
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	Prepared By and LESS THAN values are determined from Procedure 21-000-002.		
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 3.05E-01
Analyst	LESS THAN Value was Determined from Rmax.		DETECTION
SCL			LEVEL
Date Complete	RELATIVE COUNTING ERROR		3.72E-01
11/11/98	500.0%		$\mu\text{Ci/g}$
Analysis Date			
11/10/98			
Analysis Time			
09:45 PM			
Sample Point			
BY-105(2)			

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	SAC
SAMPLE.WB1 Rev. 1.0	508101ML	Date: 12/10/98

4032

WORKBOOK PAGE: SAM9

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	13	13
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	15	8
27092	COUNT TIME in MINUTES (CT)	30	30
AT OF 157	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Batch Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0164	2.0164
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.270	0.249

Batch Number	
98005701	Blank Concentration in $\mu\text{Ci/g}$ 2.05E-01
Return	Replicate Concentration in $\mu\text{Ci/g}$ < 1.89E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 2.0505E-01

Sample Prep	
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample#	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)
S98T002600	

Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
MCB	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 2.05E-01
Analyst	
SCL	LESS THAN Value was Determined from Rs.
Date Complete	
11/11/98	RELATIVE COUNTING ERROR 500.0%
Analysis Date	
11/10/98	
Analysis Time	
09:45 PM	
Sample ID	
BY-105(2)	

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	SAC
SAMPLE.WB1 Rev. 1.0	508101ML	Date: 12 Nov 98

4033

WORKBOOK PAGE: DUP10

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
WCS	DETECTOR NUMBER	13	13
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
WCS	GROSS COUNTS (GC)	7	5
27092	COUNT TIME in MINUTES (CT)	30	30
AT 0710	BACKGROUND in cpm (BKG)	0.23	0.23
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0084	2.0084
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.208	0.204
Batch Number			
98005701	Blank Concentration in $\mu\text{Ci/g}$	<	1.59E-01
Retun	Replicate Concentration in $\mu\text{Ci/g}$	<	1.56E-01
1	Maximum Concentration in $\mu\text{Ci/g}$	<	1.5889E-01
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002600			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.59E-01	DETECTION LEVEL 3.81E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
SCL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/11/98			
Analysis Date			
11/10/98			
Analysis Time			
09:45 PM			
Sample Point			
BY-105(2)			

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0	508101ML	

4034

LABCORE Completed Worklist Report for Worklist# 27093

Analyst: scl

Instrument: AB16

Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: Rerun of WL 26967, BY-105 mcb

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01	SOLID	1.80E-04	1.72E-4	95.556 %	Recovery
1 STD	0		@ALPHA01 ALPHA01E	SOLID	1.00	3.62E+00	3.620 %	Ct. Erro
2 BLNK-PREP	0		@ALPHA01 ALPHA01	SOLID	1	<3.27E-1		uCl/g
2 BLNK-PREP	0		@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	uCl/g
3 BLNK/BKG	0		@ALPHA01 ALPHA01	SOLID	1.00E+00	1.42E+00	1.420	BLNK/BKG
4 SAMPLE	S98T002845	0 F	@ALPHA01 ALPHA01	SOLID	N/A	< 3.27E-01	0.392	uCl/g
4 SAMPLE	S98T002845	0 F	@ALPHA01 ALPHA01E	SOLID	N/A	1.62E+02		% Ct. Error
5 DUP	S98T002845	0 F	@ALPHA01 ALPHA01	SOLID	<3.27E-1	2.95E-1		RPD
5 DUP	S98T002845	0 F	@ALPHA01 ALPHA01E	SOLID	1.00	1.10E+02	110.000	% Ct. Erro
6 SAMPLE	S98T002918	0 F	@ALPHA01 ALPHA01	SOLID	N/A	< 2.25E-01	0.398	uCl/g
6 SAMPLE	S98T002918	0 F	@ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02		% Ct. Error
7 DUP	S98T002918	0 F	@ALPHA01 ALPHA01	SOLID	<2.25E-1	<2.60E-1		RPD
7 DUP	S98T002918	0 F	@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	% Ct. Erro
8 SAMPLE	S98T002919	0 F	@ALPHA01 ALPHA01	SOLID	N/A	< 1.70E-01	0.401	uCl/g
8 SAMPLE	S98T002919	0 F	@ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02		% Ct. Error
9 DUP	S98T002919	0 F	@ALPHA01 ALPHA01	SOLID	<1.70E-1	<1.64E-1		RPD
9 DUP	S98T002919	0 F	@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	% Ct. Erro
10 SPK	S98T002919	0 F	@ALPHA01 ALPHA01	SOLID	4.05E-02	3.41E-02	84.198	% Recovery

Final page for worklist# 27093

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John Relyea 12/16/98
Reviewer Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27093

Analyst: S. L Instrument: AB00 16 Book# ^{W-10-98} 670 31857

Method: LA-508-101 Rev/Mod G-0

Worklist Comment: Rerun of WL 26967, BY-105 mcb

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T002845	0 F	@ALPHA01	SOLID	98000437	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
5 DUP	S98T002845	0 F	@ALPHA01	SOLID		
6 SAMPLE	S98T002918	0 F	@ALPHA01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
7 DUP	S98T002918	0 F	@ALPHA01	SOLID		
8 SAMPLE	S98T002919	0 F	@ALPHA01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
9 DUP	S98T002919	0 F	@ALPHA01	SOLID		
10 SPK	S98T002919	0 F	@ALPHA01	SOLID		

Final page for worklist # 27093

Sue Lai 11-10-98
Analyst Signature Date

Sue Z. Capron 11/11/98
Analyst Signature Date
MCS 11/11/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	16	16
STD	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	2952	3190
27093	COUNT TIME in MINUTES (CT)	30	30
AVORIT	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	98.200	106.133
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04	
98005702	Concentration in $\mu\text{Ci/L}$ =	1.65E-01	
Remun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.78E-01	
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.7153E-01	
Sample Prep			
N/A			
Sample #	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
WL27093	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
Instrument Code	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
WB27806	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
Prepared By	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
SZC			
Chemist			
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.72E-04	DETECTION LEVEL
Analyst			
SCL			
Date Complete			7.92E-07
11/11/98	RELATIVE COUNTING ERROR =	3.6%	$\mu\text{Ci/mL}$
Analysis Date			
11/10/98			
Analysis Time			
06:55 PM			
Sample Point			
BY-105 (2)			

Analyst:	SZC	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
STANDARD.WB1 Rev. 1.0		

4037

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

		BLNK-PREP	REPLICATE
Type	DETECTOR NUMBER	16	16
BLNK-PREP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	11	6
27093	COUNT TIME in MINUTES (CT)	30	30
At or by?	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.038	2.038
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.393	0.191
Batch Number			
98005702	Blank Concentration in $\mu\text{Ci/g}$	< 3.27E-01	
Re-run	Replicate Concentration in $\mu\text{Ci/g}$	< 1.59E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 3.2736E-01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
WL27093			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SZC			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 3.27E-01	DETECTION LEVEL
Analyst			
SCL	LESS THAN Value was Determined from Rmax.		
Date Complete			3.92E-01
11/11/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/10/98			
Analysis Time			
06:55 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
BLANK.WB1 Rev. 1.0		

4038

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
Type	DETECTOR NUMBER		16	16
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
Worklist	GROSS COUNTS	(GC)	11	17
27093	COUNT TIME in MINUTES	(CT)	30	30
Al or TB?	BACKGROUND in cpm	(BKG)	0.2	0.2
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.038	2.038
Matrix	EFFICIENCY FACTOR	(EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE		0.393	0.367

Batch Number	
98005702	Blank Concentration in $\mu\text{Ci/g}$ < 3.27E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ 3.05E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 3.2736E-01

Sample Prep	
FUSION01	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$
Sample /	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$
S98T002845	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By		
SZC		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 3.27E-01
Analyst		DETECTION LEVEL
SCL	LESS THAN Value was Determined from Rmax.	
Date Complete		3.92E-01
11/11/98	RELATIVE COUNTING ERROR	161.6%
Analysis Date		$\mu\text{Ci/g}$
11/10/98		
Analysis Time		
06:55 PM		
Sample Point		
BY-105 (2)		

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	SAC	
50801ML		

4039

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2
WorkList	GROSS COUNTS (GC)	14
27093	COUNT TIME in MINUTES (CT)	30
Attn: B.Y.	BACKGROUND in cpm (BKG)	0.2
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0152
Matrix	EFFICIENCY FACTOR (EFF)	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.267
Batch Number		
98005702	Blank Concentration in $\mu\text{Ci/g}$	2.24E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	3.65E-01
1	Average Concentration in $\mu\text{Ci/g}$	2.9451E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002845		
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SZC		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	2.95E-01
Analyst		
SCL		
Date Complete		
11/11/98	RELATIVE COUNTING ERROR	109.6%
Analysis Date		
11/10/98		
Analysis Time		
06:55 PM		
Sample Point		
BY-105 (2)		

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4040

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
TYPE	DETECTOR NUMBER		16	16
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
WorkList	GROSS COUNTS	(GC)	14	4
27093	COUNT TIME in MINUTES	(CT)	30	30
Alford, J. J.	BACKGROUND in cpm	(BKG)	0.2	0.2
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.0084	2.0084
Matrix	EFFICIENCY FACTOR	(EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		0.267	0.191

Batch Number	
98005702	Blank Concentration in $\mu\text{Ci/g}$ 2.25E-01
ReRun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.61E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 2.2515E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002918	

Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SZC	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 2.25E-01
Analyst	
SCL	LESS THAN Value was Determined from R_s .
Date Complete	
11/11/98	RELATIVE COUNTING ERROR 500.0%
Analysis Date	
11/10/98	
Analysis Time	
06:55 PM	
Sample Point	
BY-105 (2)	

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	<i>John Relyea</i>	SAC Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4041

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5)	(MS)	2
Work List	GROSS COUNTS	(GC)	9
27093	COUNT TIME in MINUTES	(CT)	30
Alphas B7	BACKGROUND in cpm	(BKG)	0.2
AT	SAMPLE SIZE in mL	(SS)	0.100
Test Code	DILUTION FACTOR	(DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.0384
Matrix	EFFICIENCY FACTOR	(EFF)	0.2683
SOLID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE		0.313
Batch Number			
98005702	Blank Concentration in $\mu\text{Ci/g}$	< 2.60E-01	
Repin	Replicate Concentration in $\mu\text{Ci/g}$	< 1.58E-01	
1	Maximum Concentration in $\mu\text{Ci/g}$	< 2.6039E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002918			
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SZC			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.60E-01	DETECTION LEVEL
Analyst			
SCL	LESS THAN Value was Determined from Rmax.		
Date Complete			3.92E-01
11/11/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/10/98			
Analysis Time			
06:56 PM			
Sample Point			
BY-105 (2)			

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	808101ML	

4042

WORKBOOK PAGE: SAM8

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	12	6
27093	COUNT TIME in MINUTES (CT)	30	30
Attn:IBV	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.996	1.996
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.200	0.191

Batch Number	
96005702	Blank Concentration in $\mu\text{Ci/g}$ 1.70E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.62E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 1.6991E-01

Sample Prep

FUSION01

Sample

S98T002919

Instrument Code

WB27806

Prepared By

SZC

Chemist

SAC

Analyst

SCL

Date Complete

11/11/98

Analysis Date

11/10/98

Analysis Time

06:55 PM

Sample Point

BY-105 (2)

Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$ ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$ Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$

Detection Levels and Less Than Values are determined from Procedure LA-508-002.

ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 1.70E-01DETECTION
LEVEL

LESS THAN Value was Determined from Rs.

4.01E-01
 $\mu\text{Ci/g}$

RELATIVE COUNTING ERROR

500.0%

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4043

WORKBOOK PAGE: DUP9

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Detector Number	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	6	5
27093	COUNT TIME in MINUTES (CT)	30	30
Alford By?	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.97	1.97
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.191	0.191

Batch Number	
98005702	Blank Concentration in $\mu\text{Ci/g}$ < 1.64E-01
Run	Replicate Concentration in $\mu\text{Ci/g}$ < 1.64E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 1.6400E-01

Sample Prep	
FUSION01	
Sample	
S98T002919	
Instrument Code	
WB27806	
Prepared By	
SZC	
Chemist	
SAC	
Analyst	
SCL	
Date Complete	
11/11/98	
Analysis Date	
11/10/98	
Analysis Time	
06:55 PM	
Sample Point	
BY-105 (2)	

Rs (Sample Count Rate) = $(TC / CT) - BKG$ ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$ Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$

Detection Levels and Less Than Values are determined from Procedure LA-508-002.

SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 1.64E-01	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.	
SCL		4.06E-01
Date Complete	RELATIVE COUNTING ERROR 500.0%	$\mu\text{Ci/g}$

Analyst:	SCL	Date: 11-Nov-98
Signature of Chemist:	SAC	Date: 12/Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4044

WORKBOOK PAGE: SPK10

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	61536	60362
27093	COUNT TIME in MINUTES (CT)	30	30
Attrib B?	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.996	1.996
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005702	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2683	0.2683
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.74E+03	1.71E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	<	1.6991E-01
FUSION01			
Sample #			
S98T002919	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / ((EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci}))$		
WB27806	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$		
SZC	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
Chemist			
SAC			
Analyst			
SCL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/11/98			
Analysis Date			
11/10/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.41E-02	
06:55 PM	AVG. PERCENT SPIKE RECOVERY =	84.1%	
Sample Point			
BY-105 (2)			

Analyst:	SZC	Date: 11-Nov-98
Signature of Chemist:	<i>John Palyca</i>	SAC
SPIKE.WB1 Rev. 1.0	508101ML	Date: 12/16/98

4045

LABCORE Completed Worklist Report for Worklist# 27103

Analyst: crj Instrument: AB15 Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: Rerun of WL26966,BY-105 mcb

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		0ALPHA01 ALPHA01 SOLID		1.80E-04	1.80E-4	100.000	% Recovery
1 STD	0		0ALPHA01 ALPHA01E SOLID		1.00	3.53E+00	3.530	% Ct. Error
2 BLNK-PREP	0		0ALPHA01 ALPHA01 SOLID		1	<2.28E-1		uCi/g
2 BLNK-PREP	0		0ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	uCi/g
3 BLNK/BKG	0		0ALPHA01 ALPHA01 SOLID		1.00E+00	6.98E-01	0.698	BLNK/BKG
4 SAMPLE	S98T002601	0 F	0ALPHA01 ALPHA01 SOLID		N/A <	2.28E-01	0.531	uCi/g
4 SAMPLE	S98T002601	0 F	0ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
5 DUP	S98T002601	0 F	0ALPHA01 ALPHA01 SOLID		<2.28E-1	<2.29E-1		RPD
5 DUP	S98T002601	0 F	0ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
6 SAMPLE	S98T002602	0 F	0ALPHA01 ALPHA01 SOLID		N/A <	4.62E-01	0.527	uCi/g
6 SAMPLE	S98T002602	0 F	0ALPHA01 ALPHA01E SOLID		N/A	3.40E+02		% Ct. Error
7 DUP	S98T002602	0 F	0ALPHA01 ALPHA01 SOLID		<4.62E-1	<4.32E-1		RPD
7 DUP	S98T002602	0 F	0ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
8 SPK	S98T002602	0 F	0ALPHA01 ALPHA01 SOLID		4.05E-02	3.82E-02	94.321	% Recovery
9 SAMPLE	S98T002603	0 F	0ALPHA01 ALPHA01 SOLID		N/A <	5.94E-01	0.549	uCi/g
9 SAMPLE	S98T002603	0 F	0ALPHA01 ALPHA01E SOLID		N/A	1.41E+02		% Ct. Error
10 DUP	S98T002603	0 F	0ALPHA01 ALPHA01 SOLID		<5.94E-1	4.90E-1		RPD
10 DUP	S98T002603	0 F	0ALPHA01 ALPHA01E SOLID		1.00	8.78E+01	87.800	% Ct. Error

Final page for worklist# 27103

Analyst Signature	Date	Analyst Signature	Date
	13 Nov 98		
Reviewer Signature	Date		

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27103

Analyst: CRT Instrument: AB00 15 Book# 31B57

Method: LA-508-101 Rev/Mod CT-D

Worklist Comment: Rerun of WL26966,BY-105 mcb

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T002601	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
5 DUP	S98T002601	0 F	@ALPHA01	SOLID		
6 SAMPLE	S98T002602	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
7 DUP	S98T002602	0 F	@ALPHA01	SOLID		
8 SPK	S98T002602	0 F	@ALPHA01	SOLID		
9 SAMPLE	S98T002603	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
10 DUP	S98T002603	0 F	@ALPHA01	SOLID		

Final page for worklist # 27103

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	15	15
STD	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	3348	3129
27103	COUNT TIME in MINUTES (CT)	30	30
Atom/Et	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	111.170	103.870
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04	
98005712	Concentration in $\mu\text{Ci/L}$ =	1.86E-01	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.74E-01	
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.7971E-01	
Sample Prep			
N/A			
Sample	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
WL27103	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$		
Instrument Code	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
WB26872	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
Prepared By	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
SZC			
Chemist			
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.80E-04	DETECTION LEVEL
Analyst			
CRJ			
Date Complete			1.09E-06
11/12/98	RELATIVE COUNTING ERROR =	3.5%	$\mu\text{Ci/mL}$
Analysis Date			
11/11/98			
Analysis Time			
03:45 PM			
Sample Point			
BY-105 (2)			

Analyst:	SZC	Date: 12-Nov-98
Signature of Chemist:	SAC	Date: 12/16/98
STANDARD.WB1 Rev. 1.0	508101ML	

4048

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

		BLNK-PREP	REPLICATE
Detector Number	DETECTOR NUMBER	15	15
BLNK-PREP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	11	7
27103	COUNT TIME in MINUTES (CT)	30	30
At for B?	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.064	2.064
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.279	0.279

Batch Number	
98005712	Blank Concentration in $\mu\text{Ci/g}$ < 2.28E-01
Recur	Replicate Concentration in $\mu\text{Ci/g}$ < 2.28E-01
0	Maximum Concentration in $\mu\text{Ci/g}$ < 2.2849E-01

Sample Prep	
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$

WL27103	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Instrument Code	Detection Levels and Less Than Values are determined from Procedure LA-508-002.
WB26872	

Prepared By	
SZC	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 2.28E-01
Analyst	
CRJ	LESS Than Value was Determined from Lc.
Date Complete	
11/12/98	RELATIVE COUNTING ERROR 500.0%
Analysis Date	
11/11/98	
Analysis Time	
03:45 PM	
Sample Point	
BY-105 (2)	

Analyst	CRJ	Date: 12-Nov-98
Signature of Chemist	SAC	Date: 12/1/98
BLANK.WB1 Rev. 1.0	508101ML	

4049

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
Type	DETECTOR NUMBER		15	15
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
WorkList	GROSS COUNTS	(GC)	7	6
27103	COUNT TIME in MINUTES	(CT)	30	30
Atom/B?	BACKGROUND in cpm	(BKG)	0.43	0.43
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.064	2.064
Matrix	EFFICIENCY FACTOR	(EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE		0.279	0.279

Batch Number	98005712	Blank Concentration in $\mu\text{Ci/g}$	< 2.28E-01
Rerun	1	Replicate Concentration in $\mu\text{Ci/g}$	< 2.28E-01
Sample Prep		Maximum Concentration in $\mu\text{Ci/g}$	< 2.2849E-01

FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)
S98T002601	
Instrument Code	Relative Counting Error = $\left[\frac{(\text{The Square Root of TC} + \text{BKG} * \text{CT})}{(\text{TC} - \text{BKG} * \text{CT})} \right] * 1.96 * 100$
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	SZC		
Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.28E-01
Analyst	CRJ	LESS Than Value was Determined from Lc.	DETECTION LEVEL
Date Complete	11/12/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date	11/11/98		5.31E-01 $\mu\text{Ci/g}$
Analysis Time	03:45 PM		
Sample Point	BY-105 (2)		

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist:	SAC	Date: 12/11/98
SAMPLE.WB1 Rev. 1.	508101ML	

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	7	5
27103	COUNT TIME in MINUTES (CT)	30	30
At Work By?	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.06	2.06
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.279	0.279
Batch Number			
98005712	Blank Concentration in $\mu\text{Ci/g}$	< 2.29E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 2.29E-01	
1	Maximum Concentration in $\mu\text{Ci/g}$	< 2.2894E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002601			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SZC			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.29E-01	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
CRJ			
Date Complete			5.32E-01
11/12/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/11/98			
Analysis Time			
03:45 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist:	SAC	Date: 12-Nov-98
SAMPLE.WB1 Rev. 1.	508101ML	

4051

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
Type	DETECTOR NUMBER		15	15
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
WorkList	GROSS COUNTS	(GC)	30	16
27103	COUNT TIME in MINUTES	(CT)	30	30
Alpha/B?	BACKGROUND in cpm	(BKG)	0.43	0.43
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.0816	2.0816
Matrix	EFFICIENCY FACTOR	(EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		0.570	0.399

Batch Number	
98005712	Blank Concentration in $\mu\text{Ci/g}$ 4.62E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ < 3.24E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 4.6226E-01

Sample Prep	
FUSION01	
Sample #	
S98T002602	
Instrument Code	
WB26872	
Prepared By	
SZC	
Chemist	
SAC	
Analyst	
CRJ	
Date Complete	
11/12/98	
Analysis Date	
11/11/98	
Analysis Time	
03:45 PM	
Sample Point	
BY-105 (2)	

Rs (Sample Count Rate) = $(TC / CT) - BKG$ ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$ Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$

Detection Levels and Less Than Values are determined from Procedure LA-508-002.

SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 4.62E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rs.		
Date Complete	RELATIVE COUNTING ERROR	339.9%	5.27E-01 $\mu\text{Ci/g}$

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4052

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	16	29
27103	COUNT TIME in MINUTES (CT)	30	30
Airont BY	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.096	2.096
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.399	0.537
Batch Number			
98005712	Blank Concentration in $\mu\text{Ci/g}$	< 3.21E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	4.32E-01	
1	Maximum Concentration in $\mu\text{Ci/g}$	< 4.3224E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sampler	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002602			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SZC			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 4.32E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.		
CRJ			5.23E-01
Date Complete	RELATIVE COUNTING ERROR		$\mu\text{Ci/g}$
11/12/98	500.0%		
Analysis Date			
11/11/98			
Analysis Time			
03:45 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist: <i>SA Oak</i>	SAC	Date: <i>12/11/98</i>
SAMPLE.WB1 Rev. 1. 508101ML		

4053

WORKBOOK PAGE: SPK8

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	67608	69677
27103	COUNT TIME in MINUTES (CT)	30	30
27103	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0816	2.0816
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005712	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2695	0.2695
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.83E+03	1.88E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	4.6226E-01	
FUSION01			
Sample #			
S98T002602			
Instrument Code			
WB26872			
Prepared By			
SZC			
Chemist			
SAC			
Analyst			
CRJ			
Date Complete			
11/12/98			
Analysis Date			
11/11/98			
Analysis Time			
03:45 PM			
Sample Point			
BY-105 (2)			

$$Rs \text{ (Sample Count Rate)} = (TC / CT) - BKG$$

$$\text{SAMPLE + SPIKE } \mu\text{Ci/g} = Rs * 1000\text{mL/L} * DF / ((EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci}))$$

$$QC \text{ ACTUAL} = SVal$$

$$QC \text{ FOUND} = (((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$$

$$\text{PERCENT SPIKE RECOVERY} = (QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$$

NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.

QC ACTUAL	=	4.05E-02
QC FOUND	=	3.82E-02
AVG. PERCENT SPIKE RECOVERY	=	94.3%

Analyst:	SZC	Date: 12-Nov-98
Signature of Chemist:	<i>JAC</i>	SAC Date: 12-Nov-98
SPIKE.WB1 Rev. 1.0	508101ML	

4054

WORKBOOK PAGE: SAM9

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
Type	DETECTOR NUMBER		15	15
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
Worklist	GROSS COUNTS	(GC)	34	21
27103	COUNT TIME in MINUTES	(CT)	30	30
Alfort By	BACKGROUND in cpm	(BKG)	0.43	0.43
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	1.998	1.998
Matrix	EFFICIENCY FACTOR	(EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		0.703	0.590

Batch Number		
98005712	Blank Concentration in $\mu\text{Ci/g}$	5.94E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 4.99E-01
1	Maximum Concentration in $\mu\text{Ci/g}$	< 5.9426E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sampler	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002603	
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By			
SZC			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 5.94E-01
Analyst	LESS THAN Value was Determined from Rs.		DETECTION LEVEL
CRJ			
Date Complete			5.49E-01
11/12/98	RELATIVE COUNTING ERROR	140.9%	$\mu\text{Ci/g}$
Analysis Date			
11/11/98			
Analysis Time			
03:45 PM			
Sample Point			
BY-105 (2)			

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist:		SAC Date: 12/11/98
SAMPLE.WB1 Rev. 1.	508101ML	

4055

WORKBOOK PAGE: DUP10

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	33	27
27103	COUNT TIME in MINUTES (CT)	30	30
Alfa 1000	BACKGROUND in cpm (BKG)	0.43	0.43
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9648	1.9648
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.670	0.470

Batch Number	
98005712	Blank Concentration in $\mu\text{Ci/g}$ 5.76E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ 4.04E-01
1	Average Concentration in $\mu\text{Ci/g}$ 4.8974E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample /	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002603	
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SZC	
Chemist	
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) = 4.90E-01
Analyst	
CRJ	
Date Complete	
11/12/98	RELATIVE COUNTING ERROR 87.8%
Analysis Date	
11/11/98	
Analysis Time	
03:45 PM	
Sample Point	
BY-105 (2)	

Analyst:	CRJ	Date: 12-Nov-98
Signature of Chemist:	<i>CRJ</i>	SAC Date: 13 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4056

LABCORE Completed Worklist Report for Worklist# 27132

Analyst: gll Instrument: AB15 Book#: _____

Method: LA-508-101 Rev/Mod _____

Worklist Comment: BY105, @ALPHA01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 SOLID		1.80E-04	1.61E-4	89.444 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E SOLID		1.00	3.71E+00	3.710 % Ct. Error	
2 BLNK-FREE	0		@ALPHA01 ALPHA01 SOLID		1	<2.45E-1		uCi/g
2 BLNK-FREE	0		@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	uCi/g
3 BLNK/BKG	0		@ALPHA01 ALPHA01 SOLID		1.00E+00	1.54E+00	1.540	BLNK/BKG
4 SAMPLE	S98T002712	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	1.72E-01	0.342	uCi/g
4 SAMPLE	S98T002712	0 F	@ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
5 DUP	S98T002712	0 F	@ALPHA01 ALPHA01 SOLID		<1.72E-1	<1.36E-1		RPD
5 DUP	S98T002712	0 F	@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
6 SAMPLE	S98T002739	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	2.40E-01	0.335	uCi/g
6 SAMPLE	S98T002739	0 F	@ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
7 DUP	S98T002739	0 F	@ALPHA01 ALPHA01 SOLID		<2.40E-1	<2.04E-1		RPD
7 DUP	S98T002739	0 F	@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
8 SPK	S98T002739	0 F	@ALPHA01 ALPHA01 SOLID		4.05E-02	3.48E-02	85.926 % Recovery	
9 SAMPLE	S98T002844	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	2.78E-01	0.338	uCi/g
9 SAMPLE	S98T002844	0 F	@ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
10 DUP	S98T002844	0 F	@ALPHA01 ALPHA01 SOLID		<2.78E-1	<1.34E-1		RPD
10 DUP	S98T002844	0 F	@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
11 SPK	S98T002844	0 F	@ALPHA01 ALPHA01 SOLID		4.05E-02	3.55E-02	87.654 % Recovery	

Final page for worklist# 27132

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11/16/98

LABCORE Data Entry Template for Worklist# 27132

Analyst: gll Instrument: AB00 15 Book# 31857Method: LA-508-101 Rev/Mod G-d

Worklist Comment: BY105, @ALPHA01, tdm

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@ALPHA01	SOLID		
2	BLNK-PREP				@ALPHA01	SOLID		
3	BLNK/BKG				@ALPHA01	SOLID		
4	SAMPLE	S98T002712	0	F	@ALPHA01	SOLID	98000436	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
5	DUP	S98T002712	0	F	@ALPHA01	SOLID		
6	SAMPLE	S98T002739	0	F	@ALPHA01	SOLID	98000490	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
7	DUP	S98T002739	0	F	@ALPHA01	SOLID		
8	SPK	S98T002739	0	F	@ALPHA01	SOLID		
9	SAMPLE	S98T002844	0	F	@ALPHA01	SOLID	98000437	BY-105 (2)
		Analytes Requested: ALPHA01 , ALPHA01E						
10	DUP	S98T002844	0	F	@ALPHA01	SOLID		
11	SPK	S98T002844	0	F	@ALPHA01	SOLID		

Final page for worklist # 27132

Analyst Signature gll Date 11-13-98Analyst Signature M. Brown Date 11/14/98
Reusight 11/16/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	15
STD	DISH SIZE (1, 2, or 5) (MS)	2
Worklist	GROSS COUNTS (GC)	2990
27132	COUNT TIME in MINUTES (CT)	30
Before Bkg	BACKGROUND in cpm (BKG)	0.13
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	99.537
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005741	Concentration in $\mu\text{Ci/L}$ =	1.66E-01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.66E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.6113E-01
Sample Prep		
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$	
Sample	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$	
WL27132-STD	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.61E-04
Analyst		DETECTION LEVEL
GLL		
Date Complete		6.65E-07 $\mu\text{Ci/mL}$
11/14/98	RELATIVE COUNTING ERROR =	3.7%
Analysis Date		
11/13/98		
Analysis Time		
02:10 PM		
Sample Point		
BY-105 (2)		

Analyst:	SLH2	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 11/14/98

STANDARD.WB1 Rev. 1.0

508101ML

4059

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

	BLNK-PREP	REPLICATE
Type	DETECTOR NUMBER	15
BLNK-PREP	DISH SIZE (1, 2, or 5)	2
Work List	GROSS COUNTS	5
27132	COUNT TIME in MINUTES	30
Ator 1 By	BACKGROUND in cpm	0.13
AT	SAMPLE SIZE in mL	0.100
Test Code	DILUTION FACTOR	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	1.9668
Matrix	EFFICIENCY FACTOR	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.201
Batch Number		
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 1.72E-01
Re-run	Replicate Concentration in $\mu\text{Ci/g}$	< 2.45E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.4455E-01
Sample Prep		
N/A		
Sample #	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
WL27132-BLNK	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.45E-01
Analyst		
GLL	LESS THAN Value was Determined from Rmax.	
Date Complete		
11/14/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date		
11/13/98		
Analysis Time		
02:10 PM		
Sample Point		
BY-105 (2)		

DETECTION
LEVEL3.42E-01
 $\mu\text{Ci/g}$

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 11/14/98
BLANK.WB1 Rev. 1.0	508101ML	

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2
WorkList	GROSS COUNTS	(GC)	4
27132	COUNT TIME in MINUTES	(CT)	30
AT 101 B?	BACKGROUND in cpm	(BKG)	0.13
AT	SAMPLE SIZE in mL	(SS)	0.100
Test Code	DILUTION FACTOR	(DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	1.9668
Matrix	EFFICIENCY FACTOR	(EFF)	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE		0.158
Batch Number			
98005741	Blank Concentration in $\mu\text{Ci/g}$	<	1.36E-01
Re-run	Replicate Concentration in $\mu\text{Ci/g}$	<	1.72E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	<	1.7231E-01
Sample Prep			
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)		
S98T002712			
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.72E-01	DETECTION LEVEL 3.42E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
GLL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
02:10 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0 508101ML		

4061

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	15	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	4	4
27132	COUNT TIME in MINUTES (CT)	30	30
Atop IB	BACKGROUND in cpm (BKG)	0.13	0.13
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9616	1.9616
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.158	0.158
Batch Number			
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 1.36E-01	
Repur	Replicate Concentration in $\mu\text{Ci/g}$	< 1.36E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.3691E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002712			
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.36E-01	DETECTION LEVEL 3.42E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
GLL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
02:10 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0 508101ML		

4062

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	7	3
27132	COUNT TIME in MINUTES (CT)	30	30
AT on TB?	BACKGROUND in cpm (BKG)	0.13	0.13
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0048	2.0048
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.285	0.154
Batch Number			
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 2.40E-01	
Re-run	Replicate Concentration in $\mu\text{Ci/g}$	< 1.29E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.3991E-01	
Sample Prep			
FUSION01			
Sample #	Rs (Sample Count Rate) = (TC / CT) - BKG		
S98T002739	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)		
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.40E-01	DETECTION LEVEL 3.35E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
GLL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
02:10 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.0 508101ML		

4063

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	6
27132	COUNT TIME in MINUTES (CT)	30
AT or IB?	BACKGROUND in cpm (BKG)	0.13
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0124
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.243
Batch Number		
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 2.04E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.29E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.0389E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)	
S98T002739		
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.04E-01
Analyst		
GLL	LESS THAN Value was Determined from Rmax.	
Date Complete		
11/14/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date		
11/13/98		
Analysis Time		
02:10 PM		
Sample Point		
BY-105 (2)		

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/16/98
SAMPLE.WB1 Rev. 1.0 508101ML		

4064

WORKBOOK PAGE: SPK8

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
WorkList	TOTAL COUNTS (TC)	61975	63068
27132	COUNT TIME in MINUTES (CT)	30	30
Ator/B?	BACKGROUND in cpm (BKG)	0.13	0.13
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0048	2.0048
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005741	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2695	0.2695
0	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.74E+03	1.77E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	<	2.3991E-01
FUSION01			
Sample #			
S98T002739	Rs (Sample Count Rate) = (TC / CT) - BKG		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)		
WB26872	QC ACTUAL = SVal		
Prepared By	QC FOUND = (((S+S $\mu\text{Ci/g}$ - SAMPLE $\mu\text{Ci/g}$) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))		
SLH2	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
Chemist			
SAC			
Analyst			
GLL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/14/98			
Analysis Date			
11/13/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.48E-02	
02:10 PM	AVG. PERCENT SPIKE RECOVERY =	85.9%	
Sample Point			
BY-105 (2)			

Analyst:	SLH2	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 11/14/98

SPIKE.WB1 Rev. 1.0

508101ML

4065

WORKBOOK PAGE: SAM9

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	8	3
27132	COUNT TIME in MINUTES (CT)	30	30
A for B?	BACKGROUND in cpm (BKG)	0.13	0.13
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9852	1.9852
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.326	0.154
Batch Number			
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 2.78E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.31E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.7756E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002844			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.78E-01	DETECTION LEVEL
Analyst			
GLL	LESS THAN Value was Determined from Rmax.		
Date Complete			3.38E-01
11/14/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/13/98			
Analysis Time			
02:10 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/1/98
SAMPLE.WB1 Rev. 1.0 508101ML		

4066

WORKBOOK PAGE: DUP10

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
Detector Type	DETECTOR NUMBER	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	4
27132	COUNT TIME in MINUTES (CT)	30
AT or BR	BACKGROUND in cpm (BKG)	0.13
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9944
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.158
Batch Number		
98005741	Blank Concentration in $\mu\text{Ci/g}$	< 1.34E-01
Re-run	Replicate Concentration in $\mu\text{Ci/g}$	< 1.34E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.3367E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)	
S98T002844		
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.34E-01
Analyst		
GLL	LESS THAN Value was Determined from Rmax.	
Date Complete	RELATIVE COUNTING ERROR	500.0%
11/14/98		
Analysis Date		
11/13/98		
Analysis Time		
02:10 PM		
Sample Point		
BY-105 (2)		

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/1/98
SAMPLE.WB1 Rev. 1.0	508101ML	

4067

WORKBOOK PAGE: SPK11

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
SPK	DETECTOR NUMBER	15	15
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	62670	64856
27132	COUNT TIME in MINUTES (CT)	30	30
At or IB	BACKGROUND in cpm (BKG)	0.13	0.13
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9852	1.9852
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005741	SPIKE VALUE in µCi/mL (SVal)	4.0544E-02	4.0544E-02
Regrins	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2695	0.2695
0	SAMPLE + SPIKE µCi/g (S+S)	1.78E+03	1.84E+03
Sample Prep	AVERAGE or MAXIMUM µCi/g in SAMPLE	< 2.7756E-01	
FUSION01			

Sample #
S98T002844
Instrument Code
WB26872
Prepared By
SLH2
Chemist
SAC
Analyst
GLL
Date Complete
11/14/98
Analysis Date
11/13/98
Analysis Time
02:10 PM
Sample Point
BY-105 (2)

$$Rs \text{ (Sample Count Rate)} = (TC / CT) - BKG$$

$$\text{SAMPLE} + \text{SPIKE } \mu\text{Ci/g} = Rs * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$$

QC ACTUAL = SVal

$$QC \text{ FOUND} = (((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$$

PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100

NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.

QC ACTUAL	=	4.05E-02
QC FOUND	=	3.55E-02
AVG. PERCENT SPIKE RECOVERY	=	87.6%

Analyst:	SLH2	Date: 14-Nov-98
Signature of Chemist:	<i>[Signature]</i>	SAC Date: 17/Nov-98

SPIKE.WB1 Rev. 1.0

508101ML

4068

LABCORE Completed Worklist Report for Worklist# 27133

Analyst: gll Instrument: AB18 Book#: _____

Method: LA-508-101 Rev/Mod _____

Worklist Comment: BY105, @ALPHA01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 SOLID		1.80E-04	1.64E-4	91.111 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E SOLID		1.00	3.76E+00	3.760 % Ct. Error	
2 BLNK-PREP	0		@ALPHA01 ALPHA01 SOLID		1	<1.14E-1		uCi/g
2 BLNK-PREP	0		@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	uCi/g
3 BLNK/BKG	0		@ALPHA01 ALPHA01 SOLID		1.00E+00	3.33E-01	0.333 BLNK/BKG	
4 SAMPLE	S98T002920	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	1.51E-01	0.304	uCi/g
4 SAMPLE	S98T002920	0 F	@ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
5 DUP	S98T002920	0 F	@ALPHA01 ALPHA01 SOLID		<1.51E-1	<2.23E-1		RPD
5 DUP	S98T002920	0 F	@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
6 SPK	S98T002920	0 F	@ALPHA01 ALPHA01 SOLID		4.05E-02	3.52E-02	86.914 % Recovery	
7 SAMPLE	S98T002921	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	1.88E-01	0.304	uCi/g
7 SAMPLE	S98T002921	0 F	@ALPHA01 ALPHA01E SOLID		N/A	5.00E+02		% Ct. Error
8 DUP	S98T002921	0 F	@ALPHA01 ALPHA01 SOLID		<1.88E-1	<1.15E-1		RPD
8 DUP	S98T002921	0 F	@ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000	% Ct. Error
9 SAMPLE	S98T002922	0 F	@ALPHA01 ALPHA01 SOLID		N/A <	2.93E-01	0.316	uCi/g
9 SAMPLE	S98T002922	0 F	@ALPHA01 ALPHA01E SOLID		N/A	1.55E+02		% Ct. Error
10 DUP	S98T002922	0 F	@ALPHA01 ALPHA01 SOLID		<2.93E-1	2.04E-1		RPD
10 DUP	S98T002922	0 F	@ALPHA01 ALPHA01E SOLID		1.00	1.01E+02	101.000	% Ct. Error
11 SPK	S98T002922	0 F	@ALPHA01 ALPHA01 SOLID		4.05E-02	3.43E-02	84.691 % Recovery	

Final page for worklist# 27133

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 12/20/98

LABCORE Data Entry Template for Worklist# 27133

Analyst: gll Instrument: AB00 18 Book# 31857

Method: LA-508-101 Rev/Mod G-D

Worklist Comment: BY105, @ALPHA01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T002920 0 F		@ALPHA01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ALPHA01		, ALPHA01E			
5 DUP	S98T002920 0 F		@ALPHA01	SOLID		
6 SPK	S98T002920 0 F		@ALPHA01	SOLID		
7 SAMPLE	S98T002921 0 F		@ALPHA01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ALPHA01		, ALPHA01E			
8 DUP	S98T002921 0 F		@ALPHA01	SOLID		
9 SAMPLE	S98T002922 0 F		@ALPHA01	SOLID	98000490	BY-105 (2)
	Analytes Requested: ALPHA01		, ALPHA01E			
10 DUP	S98T002922 0 F		@ALPHA01	SOLID		
11 SPK	S98T002922 0 F		@ALPHA01	SOLID		

Final page for worklist # 27133

Analyst Signature [Signature] Date 11-13-98

Analyst Signature [Signature] Date 11/14/98
[Signature] 11/16/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	18
STD	DISH SIZE (1, 2, or 5) (MS)	2
Worksheet	GROSS COUNTS (GC)	3061
27133	COUNT TIME in MINUTES (CT)	30
Analysis By	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	101.600
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005742	Concentration in $\mu\text{Ci/L}$ =	1.73E-01
Repert	Replicate Concentration in $\mu\text{Ci/L}$ =	1.55E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.6401E-01

Sample Prep	N/A	$R_s (\text{Sample Count Rate}) = (CT / CT) - BKG$
Sample	WL27133	$\text{ALPHA TOTAL } \mu\text{Ci/L} = R_s * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Instrument Code	WB27809	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Prepared By	MCB	$\text{Relative Counting Error} = [((\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT))] * 1.96 * 100$
Chemist	SAC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Analyst	GLL	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.64E-04	DETECTION LEVEL
Date Complete	11/14/98	RELATIVE COUNTING ERROR =	3.8%	
Analysis Date	11/13/98			6.14E-07 $\mu\text{Ci/mL}$
Analysis time	01:10 PM			
Sample Point	BY-105(2)			

Analyst:	MCB	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 11/16/98
STANDARD.WB1 Rev. 1.0	508101ML	

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

	BLNK-PREP	REPLICATE
Type	18	18
BLNK-PREP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	1
27133	COUNT TIME in MINUTES (CT)	30
AT for BLNK	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0416
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.135
Batch Number		
98005742	Blank Concentration in $\mu\text{Ci/g}$	< 1.14E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.14E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.135E-01
Sample Prep		
N/A		
Sample #		
WL27133-BLNK		
Instrument Code		
WB27809		
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.14E-01
Analyst		
GLL	LESS Than Value was Determined from Lc.	
Date Complete		
11/14/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date		
11/13/98		
Analysis Time		
01:10 PM		
Sample Point		
BY-105(2)		

R_s (Sample Count Rate) = $(TC / CT) - BKG$
 $ALPHA\ TOTAL\ \mu Ci/g = R_s * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/\mu Ci)$
 $Relative\ Counting\ Error = [(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
 Detection Levels and Less Than Values are determined from Procedure LA-508-002.

	DETECTION LEVEL
	3.04E-01 $\mu\text{Ci/g}$

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/14/98
BLANK.WB1 Rev. 1.0	508101ML	

4072

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	4	2
27133	COUNT TIME in MINUTES (CT)	30	30
At or From	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0416	2.0416
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.179	0.135
Batch Number			
98005742	Blank Concentration in $\mu\text{Ci/g}$	< 1.51E-01	
Run	Replicate Concentration in $\mu\text{Ci/g}$	< 1.14E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.5074E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002920			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.51E-01	DETECTION LEVEL 3.04E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
GLL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
01:10 PM			
Sample Point			
BY-105(2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/2/98
SAMPLE.WB1 Rev. 1.0 508101ML		

4073

WORKBOOK PAGE: DUPS

AT : LA-508-101 (G-0) SOLIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	1
27133	COUNT TIME in MINUTES (CT)	30
Atom/B	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0408
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.135
Batch Number		
98005742	Blank Concentration in $\mu\text{Ci/g}$	< 1.14E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 2.23E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.2344E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)	
S98T002920		
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100	
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.23E-01
Analyst	LESS Than Value was Determined from Lc.	DETECTION LEVEL
GLL		
Date Complete	RELATIVE COUNTING ERROR	500.0%
11/14/98		3.04E-01
Analysis Date		$\mu\text{Ci/g}$
11/13/98		
Analysis Time		
01:10 PM		
Sample Point		
BY-106(2)		

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 17 Nov 98
SAMPLE.WB1 Rev. 1.0 508101ML		

4074

WORKBOOK PAGE: SPK6

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	18	18
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	63352	60518
27133	COUNT TIME in MINUTES (CT)	30	30
Ator B?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0416	2.0416
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005742	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2644	0.2644
0	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.78E+03	1.70E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	<	1.5074E-01
FUSION01			
Sample #			
S98T002920	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
WB27809	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L)))$		
MCB	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
Chemist			
SAC			
Analyst			
GLL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/14/98			
Analysis Date			
11/13/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.52E-02	
01:10 PM	AVG. PERCENT SPIKE RECOVERY =	86.7%	
Sample Point			
BY-105(2)			

Analyst:	MCB	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/1/98
SPIKE.WB1 Rev. 1.0	508101ML	

4075

WORKBOOK PAGE: SAM7

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	5	8
27133	COUNT TIME in MINUTES (CT)	30	30
AT on 1B17	BACKGROUND in cpm. (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.038	2.038
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.222	0.167
Batch Number			
98005742	Blank Concentration in $\mu\text{Ci/g}$	< 1.88E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	1.41E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.8763E-01	
Sample Prep			
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	$\text{ALPHA TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002921			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.88E-01	DETECTION LEVEL 3.04E-01 $\mu\text{Ci/g}$
Analyst	LESS THAN Value was Determined from Rmax.		
GLL			
Date Complete	RELATIVE COUNTING ERROR	500.0%	
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
01:10 PM			
Sample Point			
BY-105(2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 11/14/98

SAMPLE.WB1 Rev. 1.0

508T01ML

4076

WORKBOOK PAGE: DUP8

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	18	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	3	3
27133	COUNT TIME in MINUTES (CT)	30	30
AT on IB?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.024	2.024
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.135	0.135
Batch Number			
98005742	Blank Concentration in $\mu\text{Ci/g}$	< 1.15E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.15E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.1453E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002921			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.15E-01	DETECTION LEVEL
Analyst			
GLL	LESS Than Value was Determined from Lc.		
Date Complete			3.06E-01
11/14/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/13/98			
Analysis Time			
01:10 PM			
Sample Point			
BY-105(2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 17 Nov 98
SAMPLE.WB1 Rev. 1.0 508101ML		

4077

WORKBOOK PAGE: SAM9

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	18	18
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	13	7
27133	COUNT TIME in MINUTES (CT)	30	30
Alpha B?	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9608	1.9608
Matrix	EFFICIENCY FACTOR (EFF)	0.2644	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.333	0.307
Batch Number			
98005742	Blank Concentration in $\mu\text{Ci/g}$	2.93E-01	
ReRun	Replicate Concentration in $\mu\text{Ci/g}$	< 2.70E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.9252E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG		
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)		
S98T002922			
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100		
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.93E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rs.		
GLL			
Date Complete			3.16E-01 $\mu\text{Ci/g}$
11/14/98	RELATIVE COUNTING ERROR	155.0%	
Analysis Date			
11/13/98			
Analysis Time			
01:10 PM			
Sample Point			
BY-105(2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 17 Nov 98
SAMPLE.WB1 Rev. 1.0	508101ML	

4078

WORKBOOK PAGE: DUP10

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	18
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Worklist	GROSS COUNTS (GC)	10
27133	COUNT TIME in MINUTES (CT)	30
At or By	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9668
Matrix	EFFICIENCY FACTOR (EFF)	0.2644
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.233
Batch Number		
98005742	Blank Concentration in $\mu\text{Ci/g}$	2.04E-01
Remun	Replicate Concentration in $\mu\text{Ci/g}$	2.04E-01
0	Average Concentration in $\mu\text{Ci/g}$	2.0414E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$	
S98T002922		
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27809	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	2.04E-01
Analyst		
GLL		
Date/Complete		
11/14/98	RELATIVE COUNTING ERROR	101.0%
Analysis Date		
11/13/98		
Analysis Time		
01:10 PM		
Sample Point		
BY-105(2)		

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/1/98
SAMPLE.WB1 Rev. 1.0 508101ML		

4079

WORKBOOK PAGE: SPK11

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	18	18
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	61283	59427
27133	COUNT TIME in MINUTES (CT)	30	30
Ator B7	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9608	1.9608
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005742	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0544E-02	4.0544E-02
Recur	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2644	0.2644
0	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.79E+03	1.74E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	< 2.9252E-01	
FUSION01			
Sample #			
S98T002922	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / ((EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci}))$		
WB27809	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$		
MCB	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
Chemist			
SAC			
Analyst			
GLL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
11/14/98			
Analysis Date			
11/13/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.43E-02	
01:10 PM	AVG. PERCENT SPIKE RECOVERY =	84.5%	
Sample Point			
BY-105(2)			

Analyst:	MCB	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SPIKE.WB1 Rev. 1.0		

508101ML

4080

11/14/98

19:12:02

LABCORE Completed Worklist Report for Worklist# 27134

Analyst: gll Instrument: AB16 Book#: _____

Method: LA-508-101 Rev/Mod G-0

Worklist Comment: BY105, @ALPHA01, tdm

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0ALPHA01 ALPHA01	SOLID	1.80E-04	1.73E-4	96.111 % Recovery	
1 STD	0	0ALPHA01 ALPHA01E	SOLID	1.00	3.53E+00	3.530 % Ct. Error	
2 BLNK-PREP	0	0ALPHA01 ALPHA01	SOLID	1	<2.04E-1	uCi/g	
2 BLNK-PREP	0	0ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000 uCi/g	
3 BLNK/BKG	0	0ALPHA01 ALPHA01	SOLID	1.00E+00	5.00E+00	5.000 BLNK/BKG	
4 SAMPLE	S98T002986 0 F	0ALPHA01 ALPHA01	SOLID	N/A	< 1.54E-01	0.205 uCi/g	
4 SAMPLE	S98T002986 0 F	0ALPHA01 ALPHA01E	SOLID	N/A	1.84E+02	% Ct. Error	
5 DUP	S98T002986 0 F	0ALPHA01 ALPHA01	SOLID	<1.54E-1	1.88E-1	RPD	
5 DUP	S98T002986 0 F	0ALPHA01 ALPHA01E	SOLID	1.00	1.16E+02	116.000 % Ct. Error	
6 SPK	S98T002986 0 F	0ALPHA01 ALPHA01	SOLID	4.05E-02	3.83E-02	94.568 % Recovery	
7 SAMPLE	S98T002987 0 F	0ALPHA01 ALPHA01	SOLID	N/A	< 6.62E-02	0.199 uCi/g	
7 SAMPLE	S98T002987 0 F	0ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02	% Ct. Error	
8 DUP	S98T002987 0 F	0ALPHA01 ALPHA01	SOLID	<6.62E-2	1.84E-1	RPD	
8 DUP	S98T002987 0 F	0ALPHA01 ALPHA01E	SOLID	1.00	1.01E+02	101.000 % Ct. Error	
9 SPK	S98T002987 0 F	0ALPHA01 ALPHA01	SOLID	4.05E-02	3.88E-02	95.802 % Recovery	

Final page for worklist# 27134

gll 16 Mar 99
Analyst Signature Date

_____
Analyst Signature Date

gll 10 Mar 99
Reviewer Signature Date

HNF-1663 REV. 0

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	16
STANDARD	DISH SIZE (1, 2, or 5) (MS)	2
WorkList	GROSS COUNTS (GC)	3087
27134	COUNT TIME in MINUTES (CT)	30
AT or TB	BACKGROUND in cpm (BKG)	0.03
AT	SAMPLE SIZE in mL (SS)	1.000
TestCode	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2683
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	102.870
BatchNumber	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005743	Concentration in $\mu\text{Ci/L}$ =	1.73E-01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.74E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.7344E-01
SamplePrep		
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$	
Sample#	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
WL27134-STD	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L	
InstrumentCode	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
PreparedBy		
NEW		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.73E-04
Analyst		
GLL		
DateComplete		
11/17/98	RELATIVE COUNTING ERROR =	3.5%
Analysis Date		
11/16/98		
Analysis Time		
09:20 AM		
Sample Point		
BY-105 (2)		

Analyst:	NEW	Date: 17-Nov-98
Signature of Chemist:	SAC	Date: 10 Nov 99

STANDARD.WB1 Rev. 1.0

508101ML

4082

AT : LA-508-101 (G-0)

SOLIDS

		BLANK	REPLICATE
Type	DETECTOR NUMBER	16	16
BLANK	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	8	1
27134	COUNT TIME in MINUTES (CT)	30	30
Atom/B?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.972	1.972
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.237	0.079

Batch Number	
98005743	Blank Concentration in $\mu\text{Ci/g}$ 2.04E-01
Repur	Replicate Concentration in $\mu\text{Ci/g}$ < 6.81E-02
0	Maximum Concentration in $\mu\text{Ci/g}$ < 2.0351E-01

Sample Prep	
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
WL27134-BLNK	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Instrument Code	Detection Levels and Less Than Values are determined from Procedure LA-508-002.
WB27806	

Prepared By			
NEW			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 2.04E-01
Analyst	LESS THAN Value was Determined from Rs.		DETECTION LEVEL
GLL			
Date Complete	RELATIVE COUNTING ERROR		2.05E-01 $\mu\text{Ci/g}$
11/17/98	500.0%		
Analysis Date			
11/16/98			
Analysis Time			
09:20 AM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 17-Nov-98
Signature of Chemist:	SAC	Date: 10 Nov 99

BLANK.WB1 Rev. 1.0

508101ML

4083

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	4	3
27134	COUNT TIME in MINUTES (CT)	30	30
At or by	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.972	1.972
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	0.103	0.179
Batch Number			
98005743	Blank Concentration in $\mu\text{Ci/g}$	8.89E-02	
ReRun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.54E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.5359E-01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002986			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
NEW			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.54E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rs.		
GLL			
Date Complete			2.05E-01
11/17/98	RELATIVE COUNTING ERROR	184.3%	$\mu\text{Ci/g}$
Analysis Date			
11/16/98			
Analysis Time			
09:20 AM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 17-Nov-98
Signature of Chemist:	SAC	Date: 10 Nov 98
SAMPLE.WB1 Rev. 1. 508101ML		

4084

AT : LA-508-101 (G-0)

SOLIDS

		DUPLICATE	REPLICATE
Type	DETECTOR NUMBER	16	16
DUPLICATE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	10	5
27134	COUNT TIME in MINUTES (CT)	30	30
AlforIB?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.984	1.984
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.303	0.137
Batch Number			
98005743	Blank Concentration in $\mu\text{Ci/g}$	2.59E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	1.17E-01	
0	Average Concentration in $\mu\text{Ci/g}$	1.8803E-01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002986			
Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
NEW			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	1.88E-01	DETECTION LEVEL
Analyst			
GLL			
Date Complete			2.04E-01 $\mu\text{Ci/g}$
11/17/98	RELATIVE COUNTING ERROR	116.1%	
Analysis Date			
11/16/98			
Analysis Time			
09:20 AM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 17-Nov-98
Signature of Chemist: <i>JA Cotton</i>	SAC	Date: <i>10/16/99</i>
SAMPLE.WB1 Rev. 1.	508101ML	

4085

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

Type	DETECTOR NUMBER	SPIKE	REPLICATE
SPIKE	DISH SIZE 1, 2, or 5 (MS)	16	16
Work List	TOTAL COUNTS (TC)	2	2
27134	COUNT TIME in MINUTES (CT)	67482	69427
Attn: By	BACKGROUND in cpm (BKG)	30	30
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.03	0.03
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	0.100	0.100
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	101	101
Matrix	SPIKE VOLUME in mL (SVol)	1.972	1.972
SOLID	SPIKE DILUTION FACTOR (SDF)	0.100	0.100
Batch Number	SPIKE BOOK NUMBER (Spk BN)	1	1
98005743	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	148B43	148B43
Retun	INSTRUMENT EFFICIENCY FACTOR (EFF)	4.0543E-02	4.0543E-02
0	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	0.2683	0.2683
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	1.93E+03	1.99E+03
N/A		<1.5359E-01	
Sample #			
S98T002986	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
WB27806	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$		
NEW	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
Chemist			
SAC			
Analyst			
GLL			
Date Complete			
11/17/98			
Analysis Date			
11/16/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.83E-02	
09:20 AM	AVG. PERCENT SPIKE RECOVERY =	94.5%	
Sample Point			
BY-105 (2)			

Analyst:	NEW	Date: 17-Nov-98
Signature of Chemist:	<i>SAC</i>	SAC Date: 10/14/99
SPIKE.WB1 Rev. 1.0	508101ML	

4086

AT : LA-508-101 (G-0)

SOLIDS

			SAMPLE	REPLICATE
Type	DETECTOR NUMBER		16	16
SAMPLE	DISH SIZE (1, 2, or 5)	(MS)	2	2
WorkList	GROSS COUNTS	(GC)	1	1
27134	COUNT TIME in MINUTES	(CT)	30	30
Alpha B2	BACKGROUND in cpm	(BKG)	0.03	0.03
AT	SAMPLE SIZE in mL	(SS)	0.100	0.100
Test Code	DILUTION FACTOR	(DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L)	2.0272	2.0272
Matrix	EFFICIENCY FACTOR	(EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE		0.079	0.079

Batch Number	98005743	Blank Concentration in $\mu\text{Ci/g}$	<	6.62E-02
Rerun		Replicate Concentration in $\mu\text{Ci/g}$	<	6.62E-02
0		Maximum Concentration in $\mu\text{Ci/g}$	<	6.6203E-02

Sample Prep

N/A

Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$

Sample #

ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$

S98T002987

Instrument Code

Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
 Detection Levels and Less Than Values are determined from Procedure LA-508-002.

WB27806

Prepared By

NEW

Chemist

SAC

ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) = < 6.62E-02

Analyst

GLL

LESS THAN Value was Determined from Rmax.

Date Complete

11/17/98

RELATIVE COUNTING ERROR

500.0%

DETECTION
LEVEL1.99E-01
 $\mu\text{Ci/g}$

Analysis Date

11/16/98

Analysis Time

09:20 AM

Sample Point

BY-105 (2)

Analyst:	GLL	Date: 17-Nov-98
Signature of Chemist:	SAC	Date: 11 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4087

AT : LA-508-101 (G-0)

SOLIDS

		DUPLICATE	REPLICATE
Type	DETECTOR NUMBER	16	16
DUPLICATE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	9	6
27134	COUNT TIME in MINUTES (CT)	30	30
At or B?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0312	2.0312
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	0.270	0.170
Batch Number			
98005743	Blank Concentration in $\mu\text{Ci/g}$	2.25E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	1.42E-01	
0	Average Concentration in $\mu\text{Ci/g}$	1.8366E-01	
Sample Prep			
N/A	R_s (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	$\text{ALPHA TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S987002987			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
NEW			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	1.84E-01	DETECTION LEVEL
Analyst			
GLL			
Date Complete			1.99E-01 $\mu\text{Ci/g}$
11/17/98	RELATIVE COUNTING ERROR	101.0%	
Analysis Date			
11/16/98			
Analysis Time			
09:20 AM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 17-Nov-98
Signature of Chemist: <i>GLL</i>	SAC	Date: <i>10/16/98</i>

SAMPLE.WB1 Rev. 1.

508101ML

4088

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPIKE	DISH SIZE 1, 2, or 5 (MS)	2	2
WorkList	TOTAL COUNTS (TC)	68428	70377
27134	COUNT TIME in MINUTES (CT)	30	30
Alt or BY?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0272	2.0272
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98005743	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0543E-02	4.0543E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2683	0.2683
0	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.91E+03	1.96E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	<1.5359E-01	
N/A			
Sample #			
S98T002987			
Instrument Code	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
WB27806	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
Prepared By	QC ACTUAL = SVal		
NEW	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L))))$		
Chemist	PERCENT SPIKE RECOVERY = $(QC \text{ FOUND} / QC \text{ ACTUAL}) * 100$		
SAC			
Analyst			
GLL			
Date Complete			
11/17/98			
Analysis Date			
11/16/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.88E-02	
09:20 AM	AVG. PERCENT SPIKE RECOVERY =	95.8%	
Sample Point			
BY-105 (2)			

Analyst:	NEW	Date: 17-Nov-98
Signature of Chemist:	<i>[Signature]</i>	SAC Date: <i>11 Nov 99</i>
SPIKE.WB1 Rev. 1.0	508101ML	

4089

LABCORE Completed Worklist Report for Worklist# 27135

Analyst: glt Instrument: AB16 Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: BY105, @ALPHA01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01	SOLID	1.80E-04	1.69E-4	93.88%	% Recovery
1 STD	0		@ALPHA01 ALPHA01E	SOLID	1.00	3.58E+00	3.58	% Ct. Error
2 BLNK-PREP	0		@ALPHA01 ALPHA01	SOLID	1	<2.16E-1		uCi/g
2 BLNK-PREP	0		@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	uCi/g
3 BLNK/BKG	0		@ALPHA01 ALPHA01	SOLID	1.00E+00	1.67E+00	1.670	BLNK/BKG
4 SAMPLE	S98T002988	0 F	@ALPHA01 ALPHA01	SOLID	N/A	<1.90E-01	0.307	uCi/g
4 SAMPLE	S98T002988	0 F	@ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02		% Ct. Error
5 DUP	S98T002988	0 F	@ALPHA01 ALPHA01	SOLID	<1.90E-1	<1.52E-1		RPD
5 DUP	S98T002988	0 F	@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	% Ct. Error
6 SAMPLE	S98T002989	0 F	@ALPHA01 ALPHA01	SOLID	N/A	<1.87E-01	0.302	uCi/g
6 SAMPLE	S98T002989	0 F	@ALPHA01 ALPHA01E	SOLID	N/A	5.00E+02		% Ct. Error
7 DUP	S98T002989	0 F	@ALPHA01 ALPHA01	SOLID	<1.87E-1	<2.22E-1		RPD
7 DUP	S98T002989	0 F	@ALPHA01 ALPHA01E	SOLID	1.00	5.00E+02	500.000	% Ct. Error

Final page for worklist# 27135

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 12/7/98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27135

Analyst: gh Instrument: AB00 16 Book# 31357

Method: LA-508-101 Rev/Mod G-0

Worklist Comment: BY105, @ALPHA01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T002988 0 F		@ALPHA01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
5 DUP	S98T002988 0 F		@ALPHA01	SOLID		
6 SAMPLE	S98T002989 0 F		@ALPHA01	SOLID	98000509	BY-105 (2)
	Analytes Requested: ALPHA01 , ALPHA01E					
7 DUP	S98T002989 0 F		@ALPHA01	SOLID		

Final page for worklist # 27135

Ray LaPlant 11-13-98
Analyst Signature Date

Shirley L. Uddell 11-14-98
Analyst Signature Date
MSB 11/14/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	16
STD	DISH SIZE (1, 2, or 5) (MS)	2
Worksheet	GROSS COUNTS (GC)	3043
27135	COUNT TIME in MINUTES (CT)	30
Analysis	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	31B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2683
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	101.333
Batch Number	Standard Value in $\mu\text{Ci/mL}$	1.80E-04
98005744	Concentration in $\mu\text{Ci/L}$ =	1.70E-01
ReRun	Replicate Concentration in $\mu\text{Ci/L}$ =	1.68E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	1.6895E-01
Sample Prep		
N/A		
Sample	Rs (Sample Count Rate) = (TC / CT) - BKG	
WL27135-STD	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL} * DF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
Instrument Code	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L	
WB27806	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT))] * 1.96 * 100$	
Prepared By	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
SLH2		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	1.69E-04
Analyst		
GLL		
Date Complete		
11/14/98	RELATIVE COUNTING ERROR =	3.6%
Analysis Date		
11/13/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

Analyst:	SLH2	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 17 Nov 98
STANDARD.WB1 Rev. 1.0	508101ML	

4092

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

	BLNK-PREP	REPLICATE
Type	16	16
BLNK-PREP	2	2
Work List	4	6
27135	30	30
Alpha-BK	0.1	0.1
AT	0.100	0.100
Test Code	101	101
@ALPHA01	1.9872	1.9872
Matrix	0.2683	0.2683
SOLID	0.179	0.265
Batch Number		
98005744	< 1.53E-01	
Re-run	< 2.26E-01	
0	< 2.2613E-01	
Sample Prep		
N/A		
Sample		
WL27135-BLNK		
Instrument Code		
WB27806		
Prepared By		
SLH2		
Chemist		
SAC		
Analyst		
GLL		
Date Complete		
11/14/98		
Analysis Date		
11/13/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

$$Rs \text{ (Sample Count Rate)} = (TC / CT) - BKG$$

$$ALPHA \text{ TOTAL } \mu Ci/g = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/\mu Ci)$$

$$\text{Relative Counting Error} = [(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$$

Detection Levels and Less Than Values are determined from Procedure LA-508-002.

ALPHA TOTAL in $\mu Ci/g$ (Maximum) = < 2.26E-01DETECTION
LEVEL

LESS THAN Value was Determined from Rmax.

3.07E-01
 $\mu Ci/g$

RELATIVE COUNTING ERROR

500.0%

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12/1/98
BLANK.WB1 Rev. 1.0	508101ML	

4093

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	1	5
27135	COUNT TIME in MINUTES (CT)	30	30
Attrib	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9872	1.9872
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.135	0.222

Batch Number	
98005744	Blank Concentration in $\mu\text{Ci/g}$ < 1.15E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.90E-01
0	Maximum Concentration in $\mu\text{Ci/g}$ < 1.8963E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	$ALPHA \text{ TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T002988	

Instrument Code	Relative Counting Error = $[(The \text{ Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
SLH2	
Chemist	

SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.90E-01	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
GLL			
Date/Complete	RELATIVE COUNTING ERROR	500.0%	3.07E-01 $\mu\text{Ci/g}$
11/14/98			
Analysis Date			
11/13/98			
Analysis time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Noy-98
Signature of Chemist:		SAC
SAMPLE.WB1 Rev. 1.	508101ML	Date: 11/14/98

4094

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	1	4
27135	COUNT TIME in MINUTES (CT)	30	30
Ator/Bt	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.9992	1.9992
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.135	0.179

Batch Number	
98005744	Blank Concentration in $\mu\text{Ci/g}$ < 1.14E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.52E-01
0	Maximum Concentration in $\mu\text{Ci/g}$ < 1.5170E-01

Sample Prep	
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample /	ALPHA TOTAL $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$
S98T002988	

Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 1.52E-01
Analyst	LESS Than Value was Determined from Lc.		DETECTION LEVEL
GLL			3.05E-01
Date Complete	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
11/14/98			
Analysis Date			
11/13/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:		SAC Date: 12 Nov 98
SAMPLE.WB1 Rev. 1.	508101ML	

4095

WORKBOOK PAGE: SAM6

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	4	5
27135	COUNT TIME in MINUTES (CT)	30	30
Attrib By	BACKGROUND in cpm (BKG)	0.1	0.1
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0192	2.0192
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.179	0.222
Batch Number			
98005744	Blank Concentration in $\mu\text{Ci/g}$	< 1.50E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.87E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.8663E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002989			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
SLH2			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.87E-01	DETECTION LEVEL
Analyst			
GLL	LESS THAN Value was Determined from Rmax.		
Date Complete			3.02E-01
11/14/98	RELATIVE COUNTING ERROR	500.0%	$\mu\text{Ci/g}$
Analysis Date			
11/13/98			
Analysis Time			
02:00 PM			
Sample Point			
BY-105 (2)			

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	SAC	Date: 12 Nov 98
SAMPLE.WB1 Rev. 1. 508101ML		

4096

WORKBOOK PAGE: DUP7

AT : LA-508-101 (G-0)

SOLIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	4
27135	COUNT TIME in MINUTES (CT)	30
Alpha (B)	BACKGROUND in cpm (BKG)	0.1
AT	SAMPLE SIZE in mL (SS)	0.100
Test Code	DILUTION FACTOR (DF)	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0228
Matrix	EFFICIENCY FACTOR (EFF)	0.2683
SOLID	Lc, Rmax, or Rs(SAMPLE RATE) as APPROPRIATE	0.179
Batch Number		0.265
98005744	Blank Concentration in $\mu\text{Ci/g}$	< 1.50E-01
ReRun	Replicate Concentration in $\mu\text{Ci/g}$	< 2.22E-01
0	Maximum Concentration in $\mu\text{Ci/g}$	< 2.2215E-01
Sample Prep		
FUSION01	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = Rs * 1000mL/L * DF / (EFF * SS * Dg/L * 2220000dpm/ μCi)	
S98T002989		
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100	
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.22E-01
Analyst		
GLL	LESS THAN Value was Determined from Rmax.	
Date Complete		
11/14/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date		
11/13/98		
Analysis Time		
02:00 PM		
Sample Point		
BY-105 (2)		

Analyst:	GLL	Date: 14-Nov-98
Signature of Chemist:	<i>JACott</i>	SAC
SAMPLE.WB1 Rev. 1.	508101ML	Date: 12 Nov 98

4097

LABCORE Completed Worklist Report for Worklist# 27623

Analyst: scl Instrument: AB15 Book#: _____
Method: LA-508-101 Rev/Mod _____
Worklist Comment: BY-105 FOR @ALPHA-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01	LIQUID	2.61E-04	2.45E-4	93.870	% Recovery
1 STD	0		@ALPHA01 ALPHA01E	LIQUID	1.00	2.97E+00	2.970	% Ct. Error
2 BLNK	0		@ALPHA01 ALPHA01	LIQUID	1	<7.96E-3		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01	LIQUID	1.00E+00	6.11E-01	0.611	BLNK/BKG
4 SAMPLE	S98T003383	0	@ALPHA01 ALPHA01	LIQUID	N/A	< 1.33E-02	1.90E-002	uCi/mL
4 SAMPLE	S98T003383	0	@ALPHA01 ALPHA01E	LIQUID	N/A	5.00E+02		% Ct. Error
5 DUP	S98T003383	0	@ALPHA01 ALPHA01	LIQUID	<1.33E-2	<7.96E-3		RPD
5 DUP	S98T003383	0	@ALPHA01 ALPHA01E	LIQUID	1.00	5.00E+02	500.000	% Ct. Error
6 SPK	S98T003383	0	@ALPHA01 ALPHA01	LIQUID	4.05E-02	3.83E-02	94.568	% Recovery

Final page for worklist# 27623

Analyst Signature _____ Date _____
Analyst Signature _____ Date _____
Signature *28 Dec 98*
Reviewer Signature Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27623

Analyst: S.L Instrument: AB00 Book# 45B57

Method: LA-508-101 Rev/Mod GT-0

Worklist Comment: BY-105 FOR @ALPHA-01 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	LIQUID		
2 BLNK			@ALPHA01	LIQUID		
3 BLNK/BKG			@ALPHA01	LIQUID		
4 SAMPLE	S98T003383 0		@ALPHA01	LIQUID	98000437	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
5 DUP	S98T003383 0		@ALPHA01	LIQUID		
6 SPK	S98T003383 0		@ALPHA01	LIQUID		

Final page for worklist # 27623

Sue Lai 12-21-98
Analyst Signature Date

John L. Uale 12-23-98
Analyst Signature Date
Sylvia C. Linares 12/23/98

Data Entry Comments:

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0)

LA-508-113 (B-0) STANDARD

	STANDARD	REPLICATE
Type	DETECTOR NUMBER	15
STD	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	4425
27623	COUNT TIME in MINUTES (CT)	30
At or TB ?	BACKGROUND in cpm (BKG)	0.3
AT	SAMPLE SIZE in mL (SS)	1.000
Test Code	DILUTION FACTOR (DF)	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	45B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	147.200
Batch Number	Standard Value in $\mu\text{Ci/mL}$	2.61E-04
98006239	Concentration in $\mu\text{Ci/L}$ =	2.46E-01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ =	2.43E-01
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	2.4467E-01
Sample Prep		
N/A	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF / (EFF * SS * 2220000dpm/ μCi)	
WL27623-STD	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L	
Instrument Code	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	2.45E-04
Analyst		
SCL		
Date Complete		
12/23/98	RELATIVE COUNTING ERROR =	3.0%
Analysis Date		
12/21/98		
Analysis Time		
02:50 PM		
Sample Point		
BY-105 (P)		

Analyst:	SLH2	Date: 23-Dec-98
Signature of Chemist:	SAC	Date: 28 Dec 98

STANDARD.WB1 Rev. 1.0

508101ML

4100

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0) LIQUIDS

	BLNK	REPLICATE
Type	15	15
BLNK		
DISH SIZE (1, 2, or 5)	(MS)	2
Work List	(GC)	8
27623	(CT)	30
A for TB?	(BKG)	0.3
AT	(SS)	0.500
Test Code	(DF)	10201
@ALPHA01	(DDF)	1
Matrix	(EFF)	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.233
Batch Number		
98006239	Blank Concentration in $\mu\text{Ci/L}$	< 7.96E+00
Rerun	Replicate Concentration in $\mu\text{Ci/L}$	< 7.96E+00
0	Maximum Concentration in $\mu\text{Ci/L}$	< 7.9572E+00
Sample Prep		
N/A	Rs (Sample Count Rate) = (TC / CT) - BKG	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $R_s * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
WL27623-BLNK	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 7.96E-03
Analyst		
SCL	LESS Than Value was Determined from Lc.	
Date Complete	RELATIVE COUNTING ERROR	500.0%
12/23/98		
Analysis Date		
12/21/98		
Analysis Time		
02:50 PM		
Sample Point		
BY-105 (P)		

Analyst:	SCL	Date: 23-Dec-98
Signature of Chemist:	SAC	Date: 28 Dec 98
BLANK.WB1 Rev. 1.0	508101ML	

4101

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0) LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	15	15
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	13	3
27623	COUNT TIME in MINUTES (CT)	30	30
Alt for Bk	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE SIZE in mL (SS)	0.500	0.500
Test Code	DILUTION FACTOR (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.391	0.233

Batch Number	
98006239	Blank Concentration in $\mu\text{Ci/L}$ < 1.33E+01
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 7.96E+00
0	Maximum Concentration in $\mu\text{Ci/L}$ < 1.3344E+01

Sample Prep	N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$
Sample #	S98T003383	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$
Instrument Code	WB26872	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Prepared By	SLH2	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
Chemist		Detection Levels and Less Than Values are determined from Procedure LA-508-002.

SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 1.33E-02	DETECTION LEVEL
Analyst	SCL	LESS THAN Value was Determined from Rmax.	
Date Complete	12/23/98	RELATIVE COUNTING ERROR	500.0%
Analysis Date	12/21/98		1.90E-02 $\mu\text{Ci/mL}$
Analysis Time	02:50 PM		
Sample Point	BY-105 (P)		

Analyst:	SCL	Date: 23-Dec-98
Signature of Chemist:		Date: 22 Dec 98
SAMPLE.WB1 Rev. 1.	508101ML	

4102

AT : LA-508-101 (G-0) LIQUIDS

	DUP	REPLICATE
Type	DETECTOR NUMBER	15
DUP	DISH SIZE (1, 2, or 5) (MS)	2
Work List	GROSS COUNTS (GC)	5
27623	COUNT TIME in MINUTES (CT)	30
AT For TB?	BACKGROUND in cpm (BKG)	0.3
AT	SAMPLE SIZE in mL (SS)	0.500
Test Code	DILUTION FACTOR (DF)	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2695
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.233
Batch Number	Blank Concentration in $\mu\text{Ci/L}$	< 7.96E+00
98006239	Replicate Concentration in $\mu\text{Ci/L}$	< 7.96E+00
Rerun	Maximum Concentration in $\mu\text{Ci/L}$	< 7.9572E+00
0		
Sample Prep		
N/A	Rs (Sample Count Rate) = $(TC / CT) - BKG$	
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $Rs * 1000\text{mL/L} * DF * DDF / (EFF * SS * 2220000\text{dpm}/\mu\text{Ci})$	
S98T003383	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$	
Instrument Code	Relative Counting Error = $[(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$	
WB26872	Detection Levels and Less Than Values are determined from Procedure LA-508-002.	
Prepared By		
SLH2		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 7.96E-03
Analyst	LESS Than Value was Determined from Lc.	DETECTION LEVEL
SCL		1.90E-02
Date Complete	RELATIVE COUNTING ERROR	500.0%
12/23/98		$\mu\text{Ci/mL}$
Analysis Date		
12/21/98		
Analysis Time		
02:50 PM		
Sample Point		
BY-105 (P)		

Analyst:	SCL	Date: 23-Dec-98
Signature of Chemist:	SAC	Date: 28 Dec 98
SAMPLE.WB1 - Rev. 1.	508101ML	

4103

WORKBOOK PAGE: SPK6

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	15	15
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	70623	66947
27623	COUNT TIME in MINUTES (CT)	30	30
AT Port B?	BACKGROUND in cpm (BKG)	0.3	0.3
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.500	0.500
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	10201	10201
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
LIQUID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
98006239	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0537E-02	4.0537E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2695	0.2695
0	SAMPLE + SPIKE $\mu\text{Ci/mL}$ (S+S)	8.03E+01	7.61E+01
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/mL}$ in SAMPLE <	1.3344E-02	
N/A			
Sample #			
S98T003383			
Instrument Code	Rs (Sample Count Rate) = (TC / CT) - BKG		
WB26872	SAMPLE + SPIKE $\mu\text{Ci/mL}$ = Rs * DF * DDF / (EFF * SS * 2220000dpm/ μCi)		
Prepared By	QC ACTUAL = SVal		
SLH2	QC FOUND = (((S+S $\mu\text{Ci/mL}$ - SAMPLE $\mu\text{Ci/mL}$) * ((SDF/SVol)/(DF*DDF/SS)))		
Chemist	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
SAC			
Analyst			
SCL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
12/23/98			
Analysis Date			
12/21/98	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.83E-02	
02:50 PM	AVG. PERCENT SPIKE RECOVERY =	94.5%	
Sample Point			
BY-105 (P)			

Analyst:	SLH2	Date: 23-Dec-98
Signature of Chemist:		SAC Date: 28 Dec 98
SPIKE.WB1 Rev. 1.0	508101ML	

4104

LABCORE Completed Worklist Report for Worklist# 27899

Analyst: scl

Instrument: AB16

Book#: _____

Method: LA-508-101 Rev/Mod _____

Worklist Comment: rerun of 27862, BY-105 SLH2

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		0ALPHA01 ALPHA01 SOLID		2.61E-04	2.45E-4	91.870 % Recovery	
1 STD	0		0ALPHA01 ALPHA01E SOLID		1.00	2.99E+00	2.990 % Ct. Error	
2 BLNK-PREP	0		0ALPHA01 ALPHA01 SOLID		1	<1.50E-1	uCi/g	
2 BLNK-PREP	0		0ALPHA01 ALPHA01E SOLID		1.00	1.84E+02	184.000 uCi/g	
3 BLNK/BKG	0		0ALPHA01 ALPHA01 SOLID		1.00E+00	4.44E+00	4.440 BLNK/BKG	
4 SAMPLE	S98T003393	0 F	0ALPHA01 ALPHA01 SOLID		N/A	1.10E-01	0.201 uCi/g	
4 SAMPLE	S98T003393	0 F	0ALPHA01 ALPHA01E SOLID		N/A	3.03E+02	% Ct. Error	
5 DUP	S98T003393	0 F	0ALPHA01 ALPHA01 SOLID		<1.10E-1	<1.09E-1	RPD	
5 DUP	S98T003393	0 F	0ALPHA01 ALPHA01E SOLID		1.00	5.00E+02	500.000 % Ct. Error	
6 SPK	S98T003393	0 F	0ALPHA01 ALPHA01 SOLID		4.05E-02	3.64E-02	89.877 % Recovery	
7 SAMPLE	S98T003398	0 F	0ALPHA01 ALPHA01 SOLID		N/A	3.44E-01	0.203 uCi/g	
7 SAMPLE	S98T003398	0 F	0ALPHA01 ALPHA01E SOLID		N/A	8.24E+01	% Ct. Error	
8 DUP	S98T003398	0 F	0ALPHA01 ALPHA01 SOLID		3.44E-1	3.87E-1	11.765 RPD	
8 DUP	S98T003398	0 F	0ALPHA01 ALPHA01E SOLID		1.00	6.04E+01	60.400 % Ct. Error	
9 SPK	S98T003398	0 F	0ALPHA01 ALPHA01 SOLID		4.05E-02	3.65E-02	90.123 % Recovery	
10 SAMPLE	S98T003388	0 F	0ALPHA01 ALPHA01 SOLID		N/A	1.15E-01	0.200 uCi/g	
10 SAMPLE	S98T003388	0 F	0ALPHA01 ALPHA01E SOLID		N/A	1.40E+02	% Ct. Error	
11 DUP	S98T003388	0 F	0ALPHA01 ALPHA01 SOLID		1.15E-1	<2.26E-1	RPD	
11 DUP	S98T003388	0 F	0ALPHA01 ALPHA01E SOLID		1.00	1.84E+02	184.000 % Ct. Error	
12 SPK	S98T003388	0 F	0ALPHA01 ALPHA01 SOLID		4.05E-02	3.45E-02	85.185 % Recovery	

Final page for worklist# 27899

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27899

Analyst: S. L Instrument: AB00 16 Book# 45B57

Method: LA-508-101 Rev/Mod GT-D

Worklist Comment: rerun of 27862, BY-105 SLH2

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	SOLID		
2 BLNK-PREP			@ALPHA01	SOLID		
3 BLNK/BKG			@ALPHA01	SOLID		
4 SAMPLE	S98T003393	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
5 DUP	S98T003393	0 F	@ALPHA01	SOLID		
6 SPK	S98T003393	0 F	@ALPHA01	SOLID		
7 SAMPLE	S98T003398	0 F	@ALPHA01	SOLID	98000436	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
8 DUP	S98T003398	0 F	@ALPHA01	SOLID		
9 SPK	S98T003398	0 F	@ALPHA01	SOLID		
10 SAMPLE	S98T003388	0 F	@ALPHA01	SOLID	98000490	BY-105 (2)
Analytes Requested: ALPHA01 , ALPHA01E						
11 DUP	S98T003388	0 F	@ALPHA01	SOLID		
12 SPK	S98T003388	0 F	@ALPHA01	SOLID		

Final page for worklist # 27899

Sue Lai 1-18-99
Analyst Signature Date

Shane L. Hulse 1-19-99
Analyst Signature Date

Data Entry Comments:

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	16	16
STD	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	4452	4295
27899	COUNT TIME in MINUTES (CT)	30	30
AT or B	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	45B57	45B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
LIQUID	Lc, Rmax, or Rs (SAMPLE RATE) as APPROPRIATE	148.370	143.137
Batch Number	Standard Value in $\mu\text{Ci/mL}$	2.61E-04	
99000164	Concentration in $\mu\text{Ci/L}$ =	2.49E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ =	2.40E-01	
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	2.447E-01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
WL27899-STD	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	2.45E-04	
Analysis			
SCL			
Date Complete			
01/19/99	RELATIVE COUNTING ERROR =	3.0%	
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Point			
BY-105			

Analyst:	MCB	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 18 Jan 99
STANDARD.VB1 Rev. 1.0	508101ML	

4107

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

SOLIDS

		BLNK-PREP	REPLICATE
Type	DETECTOR NUMBER	16	16
BLNK-PREP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	5	3
27899	COUNT TIME in MINUTES (CT)	30	30
At or FB	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.014	2.014
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.137	0.179
Batch Number			
99000164	Blank Concentration in $\mu\text{Ci/g}$	1.15E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	< 1.50E-01	
0	Maximum Concentration in $\mu\text{Ci/g}$	< 1.5039E-01	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
WL27899-BLNK			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } \text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.50E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rs.		
SCL			2.01E-01
Date Complete	RELATIVE COUNTING ERROR	184.3%	$\mu\text{Ci/g}$
01/19/99			
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 18 Jan 99
BLANK.WB1 Rev. 1.0 508101ML		

4108

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0) SOLIDS

	SAMPLE	REPLICATE
Type	16	16
SAMPLE	DISH SIZE (1, 2, or 5)	(MS) 2 2
Work List	GROSS COUNTS	(GC) 4 2
27899	COUNT TIME in MINUTES	(CT) 30 30
AT or TB	BACKGROUND in cpm	(BKG) 0.03 0.03
AT	SAMPLE SIZE in mL	(SS) 0.100 0.100
Test Code	DILUTION FACTOR	(DF) 101 101
@ALPHA01	DIGEST GRAMS of SOLIDS / L	(Dg/L) 2.014 2.014
Matrix	EFFICIENCY FACTOR	(EFF) 0.2683 0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.103 0.130

Batch Number	Blank Concentration in $\mu\text{Ci/g}$	8.70E-02
99000164	Replicate Concentration in $\mu\text{Ci/g}$	< 1.10E-01
Rerun	Maximum Concentration in $\mu\text{Ci/g}$	< 1.0973E-01
1		

Sample Prep	
FUSION01	$\text{Rs (Sample Count Rate) = (TC / CT) - BKG}$
Sample #	$\text{ALPHA TOTAL } \mu\text{Ci/g} = \text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$
S98T003393	

Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 1.10E-01
Analyst		
SCL	LESS THAN Value was Determined from Rs.	
Date Complete		
01/19/99	RELATIVE COUNTING ERROR	303.4%
Analysis Date		
01/18/99		
Analysis Time		
10:50 AM		
Sample Point		
BY-105		

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:		SAC Date: 19 Jan 99
SAMPLE.WB1 Rev. 1.	508101ML	

4109

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0) SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	2	2
27899	COUNT TIME in MINUTES (CT)	30	30
AT of TB	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.0312	2.0312
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.130	0.130

Batch Number	
99000164	Blank Concentration in $\mu\text{Ci/g}$ < 1.09E-01
Rerun	Replicate Concentration in $\mu\text{Ci/g}$ < 1.09E-01
1	Maximum Concentration in $\mu\text{Ci/g}$ < 1.0880E-01

Sample Prep	
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$
Sample	$ALPHA\ TOTAL\ \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$
S98T003393	

Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By	
MCB	
Chemist	

SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum)	=	< 1.09E-01	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.			
SCL				
Date Complete	RELATIVE COUNTING ERROR			1.99E-01 $\mu\text{Ci/g}$
01/19/99	500.0%			
Analysis Date				
01/18/99				
Analysis Time				
10:50 AM				
Sample Point				
BY-105				

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19 Jan 99
SAMPLE.WB1 Rev. 1.	508101ML	

4110

WORKBOOK PAGE: SPK6

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	64223	65956
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.014	2.014
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
99000164	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0532E-02	4.0532E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2683	0.2683
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.80E+03	1.85E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	1.0973E-01	
FUSION01			
Sample #			
S98T003393	R_s (Sample Count Rate) = (TC / CT) - BKG		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
WB27806	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000))/(DF/SS/Dg/L)))$		
MCB	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
Chemist			
SAC			
Analyst			
SCL			
Date Complete	NOTE: Original Sample result was a LESS THAN value. Zero (0) was subtracted from the spiked value for QC found calculation.		
01/19/99			
Analysis Date	QC ACTUAL =	4.05E-02	
01/18/99	QC FOUND =	3.64E-02	
Analysis Time	AVG. PERCENT SPIKE RECOVERY =	89.9%	
10:50 AM			
Sample Point			
BY-105			

Analyst:	MCB	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19 Jan 99
SPIKE.VB1 Rev. 1.0	508101ML	

WORKBOOK PAGE: SAM7

AT : LA-508-101 (G-0)

SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	18	8
27899	COUNT TIME in MINUTES (CT)	30	30
AT or B	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.986	1.986
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	0.570	0.237
Batch Number			
99000164	Blank Concentration in $\mu\text{Ci/g}$	4.87E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	2.02E-01	
1	Average Concentration in $\mu\text{Ci/g}$	3.4438E-01	
Sample Prep			
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample #	$ALPHA\ TOTAL\ \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003398			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	3.44E-01	DETECTION LEVEL
Analyst			
SCL			
Date Complete			2.03E-01
01/19/99	RELATIVE COUNTING ERROR	82.4%	$\mu\text{Ci/g}$
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Points			
BY-105			

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19 Jan 99
SAMPLE.WB1 Rev. 1.	508101ML	

4112

WORKBOOK PAGE: DUP8

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	16	13
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.984	1.984
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	0.503	0.403
Batch Number			
99000164	Blank Concentration in $\mu\text{Ci/g}$	4.30E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	3.45E-01	
1	Average Concentration in $\mu\text{Ci/g}$	3.8746E-01	
Sample Prep			
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	$\text{ALPHA TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003398			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	3.87E-01	DETECTION LEVEL
Analyst			
SCL			
Date Complete			2.04E-01
01/19/99	RELATIVE COUNTING ERROR	60.4%	$\mu\text{Ci/g}$
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19-Jan-99
SAMPLE.WB1 Rev. 1.	508101ML	

4113

WORKBOOK PAGE: SPK9

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	64803	65731
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	1.986	1.986
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
99000164	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0532E-02	4.0532E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2683	0.2683
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.84E+03	1.87E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	3.4438E-01	
FUSION01			
Sample			
S98T003398	R_s (Sample Count Rate) = (TC / CT) - BKG		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / ((EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci}))$		
WB27806	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000)))/(DF/SS/Dg/L)))$		
MCB	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
Chemist			
SAC			
Analyst			
SCL			
Date Complete			
01/19/99			
Analysis Date			
01/18/99	QC ACTUAL	=	4.05E-02
Analysis Time	QC FOUND	=	3.65E-02
10:50 AM	AVG. PERCENT SPIKE RECOVERY	=	90.1%
Sample Point			
BY-105			

Analyst:	MCB	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19 Jan 99
SPIKE.WB1 Rev. 1.0 508101ML		

4114

WORKBOOK PAGE: SAM10

AT : LA-508-101 (G-0) SOLIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	16	16
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	6	4
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.022	2.022
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.170	0.103
Batch Number			
99000164	Blank Concentration in $\mu\text{Ci/g}$	1.43E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	8.67E-02	
1	Average Concentration in $\mu\text{Ci/g}$	1.1461E-01	
Sample Prep			
FUSION01	Rs (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	ALPHA TOTAL $\mu\text{Ci/g}$ = $Rs * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
398T003388			
Instrument Code	Relative Counting Error = $[(The\ Square\ Root\ of\ TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Average) =	1.15E-01	DETECTION LEVEL
Analyst			
SCL			
Date Complete			2.00E-01
01/19/99	RELATIVE COUNTING ERROR	140.0%	$\mu\text{Ci/g}$
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: Jan 99

SAMPLE.WB1 Rev. 1.

508101ML

4115

WORKBOOK PAGE: DUP11

AT : LA-508-101 (G-0)

SOLIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	16	16
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	3	9
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE SIZE in mL (SS)	0.100	0.100
Test Code	DILUTION FACTOR (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.024	2.024
Matrix	EFFICIENCY FACTOR (EFF)	0.2683	0.2683
SOLID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.179	0.270
Batch Number			
99000164	Blank Concentration in $\mu\text{Ci/g}$	< 1.50E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/g}$	2.26E-01	
1	Maximum Concentration in $\mu\text{Ci/g}$	< 2.2620E-01	
Sample Prep			
FUSION01	R_s (Sample Count Rate) = $(TC / CT) - BKG$		
Sample	$\text{ALPHA TOTAL } \mu\text{Ci/g} = R_s * 1000\text{mL/L} * DF / (EFF * SS * Dg/L * 2220000\text{dpm}/\mu\text{Ci})$		
S98T003388			
Instrument Code	Relative Counting Error = $[(\text{The Square Root of } TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100$		
WB27806	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/g}$ (Maximum) =	< 2.26E-01	DETECTION LEVEL
Analyst			
SCL	LESS THAN Value was Determined from Rmax.		
Date Complete			2.00E-01
01/19/99	RELATIVE COUNTING ERROR	184.3%	$\mu\text{Ci/g}$
Analysis Date			
01/18/99			
Analysis Time			
10:50 AM			
Sample Point			
BY-105			

Analyst:	SCL	Date: 19-Jan-99
Signature of Chemist:	SAC	Date: 19 Jan 99
SAMPLE.WB1 Rev. 1.	508101ML	

4116

WORKBOOK PAGE: SPK12

AT : LA-508-101 (G-0) LA-508-113 (B-0) SPIKED SAMPLE

		SPIKE	REPLICATE
Type	DETECTOR NUMBER	16	16
SPK	DISH SIZE 1, 2, or 5 (MS)	2	2
Work List	TOTAL COUNTS (TC)	66021	57136
27899	COUNT TIME in MINUTES (CT)	30	30
AT or TB?	BACKGROUND in cpm (BKG)	0.03	0.03
AT	SAMPLE VOLUME in mL (Spiked Vial) (SS)	0.100	0.100
Test Code	SAMPLE DILUTION FACTOR (Spiked Vial) (DF)	101	101
@ALPHA01	DIGEST GRAMS of SOLIDS / L (Dg/L)	2.022	2.022
Matrix	SPIKE VOLUME in mL (SVol)	0.100	0.100
SOLID	SPIKE DILUTION FACTOR (SDF)	1	1
Batch Number	SPIKE BOOK NUMBER (Spk BN)	148B43	148B43
99000164	SPIKE VALUE in $\mu\text{Ci/mL}$ (SVal)	4.0532E-02	4.0532E-02
Rerun	INSTRUMENT EFFICIENCY FACTOR (EFF)	0.2683	0.2683
1	SAMPLE + SPIKE $\mu\text{Ci/g}$ (S+S)	1.85E+03	1.60E+03
Sample Prep	AVERAGE or MAXIMUM $\mu\text{Ci/g}$ in SAMPLE	1.1461E-01	
FUSION01			
Sample			
S98T003388	R_s (Sample Count Rate) = (TC / CT) - BKG		
Instrument Code	SAMPLE + SPIKE $\mu\text{Ci/g}$ = $R_s * 1000\text{mL/L} * DF / (EFF * SS * \text{Dg/L} * 2220000\text{dpm}/\mu\text{Ci})$		
WB27806	QC ACTUAL = SVal		
Prepared By	QC FOUND = $((S+S \mu\text{Ci/g} - \text{SAMPLE } \mu\text{Ci/g}) * ((SDF/(SVol*1000)))/(DF/SS/Dg/L)))$		
MCB	PERCENT SPIKE RECOVERY = (QC FOUND / QC ACTUAL) * 100		
Chemist			
SAC			
Analyst			
SCL			
Date Complete			
01/19/99			
Analysis Date			
01/18/99	QC ACTUAL =	4.05E-02	
Analysis Time	QC FOUND =	3.45E-02	
10:50 AM	AVG. PERCENT SPIKE RECOVERY =	85.0%	
Sample Point			
BY-105			

Analyst	MCB	Date: 19-Jan-99
Signature of Chemist	SAC	Date: 01-19-99
SPIKE.WB1 Rev. 1.0 508101ML		

4117

LABCORE Completed Worklist Report for Worklist# 28057

Analyst: crj Instrument: AB13 Book#: _____

Method: LA-508-101 Rev/Mod _____

Worklist Comment: BY105 FIELD BLANK FOR @ALPHA01

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@ALPHA01 ALPHA01 LIQUID		2.61E-04	2.33E-4	89.272 % Recovery	
1 STD	0		@ALPHA01 ALPHA01E LIQUID		1.00	2.89E+00	2.890 % Ct. Error	
2 BLNK	0		@ALPHA01 ALPHA01 LIQUID		1	<5.78E-8		uCi/mL
2 BLNK	0		@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	uCi/mL
3 BLNK/BKG	0		@ALPHA01 ALPHA01 LIQUID		1.00E+00	8.33E-01	0.833	BLNK/BKG
4 SAMPLE	S98T002391	0	@ALPHA01 ALPHA01 LIQUID		N/A <	7.02E-08	1.43e-007	uCi/mL
4 SAMPLE	S98T002391	0	@ALPHA01 ALPHA01E LIQUID		N/A	5.00E+02		% Ct. Error
5 DUP	S98T002391	0	@ALPHA01 ALPHA01 LIQUID		<7.02E-8	<8.26E-8		RPD
5 DUP	S98T002391	0	@ALPHA01 ALPHA01E LIQUID		1.00	5.00E+02	500.000	% Ct. Error

Final page for worklist# 28057

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John Delaney
Reviewer Signature _____ Date 29 Jan 99

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 28057

Analyst: CRT Instrument: AB00 13 Book# 45B57

Method: LA-508-101 Rev/Mod G-0

Worklist Comment: BY105 FIELD BLANK FOR @ALPHA01

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@ALPHA01	LIQUID		
2 BLNK			@ALPHA01	LIQUID		
3 BLNK/BKG			@ALPHA01	LIQUID		
4 SAMPLE	S98T002391 0		@ALPHA01	LIQUID	98000436	BY-105 (2)
			Analytes Requested: ALPHA01 , ALPHA01E			
5 DUP	S98T002391 0		@ALPHA01	LIQUID		

Final page for worklist # 28057

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

AT : LA-508-101 (G-0) LA-508-113 (B-0) STANDARD

		STANDARD	REPLICATE
Type	DETECTOR NUMBER	13	13
STD	DISH SIZE (1, 2, or 5) (MS)	2	2
Work List	GROSS COUNTS (GC)	4607	4609
28057	COUNT TIME in MINUTES (CT)	30	30
AT or TB ?	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	1.000	1.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	STANDARD BOOK NUMBER (Std BN)	45B57	45B57
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
LIQUID	Lc, Rmax, or Rs,(SAMPLE RATE) as APPROPRIATE	153.367	153.433
Batch Number	Standard Value in $\mu\text{Ci/mL}$	2.61E-04	
99000329	Concentration in $\mu\text{Ci/L}$ =	2.33E-01	
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ =	2.33E-01	
0	AVERAGE CONCENTRATION in $\mu\text{Ci/L}$ =	2.3258E-01	

Sample Prep	
N/A	$\text{Rs (Sample Count Rate)} = (\text{TC} / \text{CT}) - \text{BKG}$
Sample #	$\text{ALPHA TOTAL } \mu\text{Ci/L} = \text{Rs} * 1000\text{mL/L} * \text{DF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$
WL28057-STD	$\text{ALPHA TOTAL } \mu\text{Ci/mL} = \text{ALPHA TOTAL } \mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	$\text{Relative Counting Error} = [(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Prepared By		
MCB		
Chemist		
SAC	ALPHA TOTAL CONCENTRATION in $\mu\text{Ci/mL}$ =	2.33E-04
Analyst		
CRJ		
Date Complete		
01/29/99	RELATIVE COUNTING ERROR =	2.9%
Analysis Date		
01/28/99		
Analysis Time		
02:00 PM		
Sample Point		
BY-105		

Analyst:	MCB	Date: 29-Jan-99
Signature of Chemist:	<i>John Relyea</i>	SAC Date: 29 Jan 99
STANDARD.WB1 Rev. 1.0	508101ML	

4120

WORKBOOK PAGE: BLANK2

AT : LA-508-101 (G-0)

LIQUIDS

		BLNK	REPLICATE
Type	DETECTOR NUMBER	13	13
BLNK	DISH SIZE (1, 2, or 5) (MS)	2	2
WorkList	GROSS COUNTS (GC)	5	5
28057	COUNT TIME in MINUTES (CT)	30	30
Alpha Bk	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	5.000	5.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
LIQUID	Lc, Rmax, or Rs, (SAMPLE RATE) as APPROPRIATE	0.191	0.191

Batch Number	
99000329	Blank Concentration in $\mu\text{Ci/L}$ < 5.78E-05
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 5.78E-05
0	Maximum Concentration in $\mu\text{Ci/L}$ < 5.7773E-05

Sample Prep	N/A	Rs (Sample Count Rate) = (TC / CT) - BKG
Sample	WL28057-BLNK	ALPHA TOTAL $\mu\text{Ci/L}$ = Rs * 1000mL/L * DF * DDF / (EFF * SS * 2220000dpm/ μCi)
Instrument Code	WB27810	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L}$ / 1000mL/L
Prepared By	MCB	Relative Counting Error = [(The Square Root of TC + BKG * CT) / (TC - BKG * CT)] * 1.96 * 100
Chemist		Detection Levels and Less Than Values are determined from Procedure LA-508-002.

SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 5.78E-08	DETECTION LEVEL
Analyst	LESS Than Value was Determined from Lc.		
CRJ			
Date Complete	01/29/99	RELATIVE COUNTING ERROR	500.0%

Analysis Date	01/28/99
Analysis Time	02:00 PM
Sample Point	BY-105

Analyst:	CRJ	Date: 29-Jan-99
Signature of Chemist:	SAC	Date: 29 Jan 99
BLANK.WB1 Rev. 1.0	508101ML	

4121

WORKBOOK PAGE: SAM4

AT : LA-508-101 (G-0)

LIQUIDS

		SAMPLE	REPLICATE
Type	DETECTOR NUMBER	13	13
SAMPLE	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	7	7
28057	COUNT TIME in MINUTES (CT)	30	30
Atto/Bg	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	5.000	5.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	0.232	0.232
Batch Number			
99000329	Blank Concentration in $\mu\text{Ci/L}$	< 7.02E-05	
ReRun	Replicate Concentration in $\mu\text{Ci/L}$	< 7.02E-05	
0	Maximum Concentration in $\mu\text{Ci/L}$	< 7.0240E-05	
Sample Prep			
N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$		
Sample #	ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$		
S98T002391	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$		
Instrument Code	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$		
WB27810	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Prepared By			
MCB			
Chemist			
SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 7.02E-08	DETECTION LEVEL
Analyst	LESS THAN Value was Determined from Rmax.		
CRJ			
Date Complete	RELATIVE COUNTING ERROR	500.0%	1.43E-07 $\mu\text{Ci/mL}$
01/29/99			
Analysis Date			
01/28/99			
Analysis Time			
02:00 PM			
Sample Point			
BY-105			

Analyst:	CRJ	Date: 29-Jan-99
Signature of Chemist:	SAC	Date: 29 Jan 99
SAMPLE.WB1 Rev. 1.0 508101MX		

4122

WORKBOOK PAGE: DUP5

AT : LA-508-101 (G-0)

LIQUIDS

		DUP	REPLICATE
Type	DETECTOR NUMBER	13	13
DUP	DISH SIZE (1, 2, or 5) (MS)	2	2
Worklist	GROSS COUNTS (GC)	8	6
28057	COUNT TIME in MINUTES (CT)	30	30
Alton B7	BACKGROUND in cpm (BKG)	0.2	0.2
AT	SAMPLE SIZE in mL (SS)	5.000	5.000
Test Code	DILUTION FACTOR (DF)	1	1
@ALPHA01	DIGEST DILUTION FACTOR (DDF)	1	1
Matrix	EFFICIENCY FACTOR (EFF)	0.2971	0.2971
LIQUID	Lc, Rmax, or Rs.(SAMPLE RATE) as APPROPRIATE	0.272	0.191

Batch Number	
99000329	Blank Concentration in $\mu\text{Ci/L}$ < 8.26E-05
Rerun	Replicate Concentration in $\mu\text{Ci/L}$ < 5.78E-05
0	Maximum Concentration in $\mu\text{Ci/L}$ < 8.2618E-05

Sample Prep	N/A	Rs (Sample Count Rate) = $(\text{TC} / \text{CT}) - \text{BKG}$
NA		ALPHA TOTAL $\mu\text{Ci/L}$ = $\text{Rs} * 1000\text{mL/L} * \text{DF} * \text{DDF} / (\text{EFF} * \text{SS} * 2220000\text{dpm}/\mu\text{Ci})$
Sample #	S98T002391	ALPHA TOTAL $\mu\text{Ci/mL}$ = ALPHA TOTAL $\mu\text{Ci/L} / 1000\text{mL/L}$
Instrument Code	WB27810	Relative Counting Error = $[(\text{The Square Root of TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})] * 1.96 * 100$
Prepared By	MCB	Detection Levels and Less Than Values are determined from Procedure LA-508-002.

Chemist	SAC	ALPHA TOTAL in $\mu\text{Ci/mL}$ (Maximum) =	< 8.26E-08	DETECTION LEVEL
Analyst	CRJ	LESS THAN Value was Determined from Rmax.		
Date Complete	01/29/99	RELATIVE COUNTING ERROR	500.0%	1.43E-07 $\mu\text{Ci/mL}$
Analysis Date	01/28/99			
Analysis Time	02:00 PM			
Sample Point	BY-105			

Analyst:	CRJ	Date: 29-Jan-99
Signature of Chemist:	SAC	Date: 29 Jan 99
SAMPLE.WB1 Rev. 1.0	508101MLC	

4123

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 14 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 4124 THRU 4485)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	GAMMA ENERGY ANALYSIS, GEA WORKLIST 26735 THRU 27142		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of 17
Continued

See 4 of 17

See 5 of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 7
of 17
Cont.

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26735

Analyst: gll

Instrument: GEA02

Book#: 28B57

Method: LA-548-121

Rev/Mod F-φ

Worklist Comment: BY-105 FOR @GEA-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@GEA-01 C060-02 LIQUID		4.531e-03	4.82e-03	106.378 % Recovery	
1 STD	0	@GEA-01 C060-02E LIQUID		1	3.11	3.110 % Ct Error	
1 STD	0	@GEA-01 CS13702 LIQUID		7.453e-03	7.71e-03	103.448 % Recovery	
1 STD	0	@GEA-01 CS13702E LIQUID		1	3.00	3.000 % Ct Error	
2 BLNK	0	@GEA-01 C060-02 LIQUID		1	<1.67e-5	uCi/mL	
2 BLNK	0	@GEA-01 CS13702 LIQUID		1	<1.57e-5	uCi/mL	
3 SAMPLE	S98T002392 0	@GEA-01 C060-02 LIQUID		N/A	1.886e-05	1.886e-05 uCi/mL	
3 SAMPLE	S98T002392 0	@GEA-01 C060-02E LIQUID		N/A	n/a	% Ct. Error	
3 SAMPLE	S98T002392 0	@GEA-01 CS13702 LIQUID		N/A	3.829e-05	3.83e-05 uCi/mL	
3 SAMPLE	S98T002392 0	@GEA-01 CS13702E LIQUID		N/A	n/a	% Ct. Error	
4 DUP	S98T002392 0	@GEA-01 C060-02 LIQUID		<1.99e-5	<1.88e-5	RPD	
4 DUP	S98T002392 0	@GEA-01 C060-02E LIQUID		1	n/a	% Ct Error	
4 DUP	S98T002392 0	@GEA-01 CS13702 LIQUID		<3.83e-5	<4.21e-5	RPD	
4 DUP	S98T002392 0	@GEA-01 CS13702E LIQUID		1	n/a	% Ct Error	
5 SAMPLE	S98T002814 0	@GEA-01 C060-02 LIQUID		N/A	3.096e-03	3.10e-03 uCi/mL	
5 SAMPLE	S98T002814 0	@GEA-01 C060-02E LIQUID		N/A	n/a	% Ct. Error	
5 SAMPLE	S98T002814 0	@GEA-01 CS13702 LIQUID		N/A	1.620e+02	uCi/mL	
5 SAMPLE	S98T002814 0	@GEA-01 CS13702E LIQUID		N/A	0.140	% Ct. Error	
6 DUP	S98T002814 0	@GEA-01 C060-02 LIQUID		<3.10e-3	<3.10e-3	RPD	
6 DUP	S98T002814 0	@GEA-01 C060-02E LIQUID		1	n/a	% Ct Error	
6 DUP	S98T002814 0	@GEA-01 CS13702 LIQUID		1.62e+02	1.62e+02	0.000 RPD	
6 DUP	S98T002814 0	@GEA-01 CS13702E LIQUID		1	0.140	% Ct Error	
7 SAMPLE	S98T002815 0	@GEA-01 C060-02 LIQUID		N/A	3.430e-03	3.43e-03 uCi/mL	
7 SAMPLE	S98T002815 0	@GEA-01 C060-02E LIQUID		N/A	n/a	% Ct. Error	
7 SAMPLE	S98T002815 0	@GEA-01 CS13702 LIQUID		N/A	1.770e+02	uCi/mL	
7 SAMPLE	S98T002815 0	@GEA-01 CS13702E LIQUID		N/A	0.140	% Ct. Error	
8 DUP	S98T002815 0	@GEA-01 C060-02 LIQUID		<3.43e-3	<3.33e-3	RPD	
8 DUP	S98T002815 0	@GEA-01 C060-02E LIQUID		1	n/a	% Ct Error	
8 DUP	S98T002815 0	@GEA-01 CS13702 LIQUID		1.77e+02	1.77e+02	0.000 RPD	
8 DUP	S98T002815 0	@GEA-01 CS13702E LIQUID		1	0.140	% Ct Error	
9 SAMPLE	S98T002816 0	@GEA-01 C060-02 LIQUID		N/A	3.421e-03	3.42e-03 uCi/mL	
9 SAMPLE	S98T002816 0	@GEA-01 C060-02E LIQUID		N/A	n/a	% Ct. Error	
9 SAMPLE	S98T002816 0	@GEA-01 CS13702 LIQUID		N/A	1.980e+02	uCi/mL	
9 SAMPLE	S98T002816 0	@GEA-01 CS13702E LIQUID		N/A	0.130	% Ct. Error	
10 DUP	S98T002816 0	@GEA-01 C060-02 LIQUID		<3.42e-3	<3.15e-3	RPD	
10 DUP	S98T002816 0	@GEA-01 C060-02E LIQUID		1	n/a	% Ct Error	
10 DUP	S98T002816 0	@GEA-01 CS13702 LIQUID		1.98e+02	1.98e+02	0.000 RPD	
10 DUP	S98T002816 0	@GEA-01 CS13702E LIQUID		1	0.130	% Ct Error	
11 SAMPLE	S98T002817 0	@GEA-01 C060-02 LIQUID		N/A	3.565e-03	3.57e-03 uCi/mL	
11 SAMPLE	S98T002817 0	@GEA-01 C060-02E LIQUID		N/A	n/a	% Ct. Error	
11 SAMPLE	S98T002817 0	@GEA-01 CS13702 LIQUID		N/A	1.900e+02	uCi/mL	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26735

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
11 SAMPLE	S98T002817	0	@GEA-01	CS13702E LIQUID	N/A	0.130		% Ct. Error
12 DUP	S98T002817	0	@GEA-01	CO60-02 LIQUID	<3.57e-3	<3.30e-3		RPD
12 DUP	S98T002817	0	@GEA-01	CO60-02E LIQUID	1	n/a		% Ct. Error
12 DUP	S98T002817	0	@GEA-01	CS13702 LIQUID	1.90e+02	1.91e+02	0.525	RPD
12 DUP	S98T002817	0	@GEA-01	CS13702E LIQUID	1	0.130	0.130	% Ct. Error

Comments Section:

Comments for sample# S98T002814 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002815 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002816 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002817 and test @GEA-01 .
DL=0 => n/a.

Final page for worklist# 26735

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11-16-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26735

Analyst: gll Instrument: GEA00 2 Book# 28857

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY-105 FOR @GEA-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	LIQUID		
2 BLNK			@GEA-01	LIQUID		
~~3 SAMPLE~~	~~S98T002392 0~~	~~@GEA-01~~	~~LIQUID~~	~~98000436~~	~~BY-105 (2)~~	~~Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E~~
~~4 DUP~~	~~S98T002392 0~~	~~@GEA-01~~	~~LIQUID~~			
5 SAMPLE	S98T002814 0	@GEA-01	LIQUID	98000437	BY-105 (2)	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E
6 DUP	S98T002814 0	@GEA-01	LIQUID			
7 SAMPLE	S98T002815 0	@GEA-01	LIQUID	98000437	BY-105 (2)	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E
8 DUP	S98T002815 0	@GEA-01	LIQUID			
9 SAMPLE	S98T002816 0	@GEA-01	LIQUID	98000437	BY-105 (2)	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E
10 DUP	S98T002816 0	@GEA-01	LIQUID			
11 SAMPLE	S98T002817 0	@GEA-01	LIQUID	98000437	BY-105 (2)	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E
12 DUP	S98T002817 0	@GEA-01	LIQUID			

Seen 3/24/99

Final page for worklist # 26735

gll 11/11/98
Analyst Signature Date

MEB 11/14/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****

* 222-S Laboratory Counting Room 11-NOV-1998 22:25:53.68 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: WL26735-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Newright

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2798.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:21.25 sec
 Dead Time: 0.7%

Verified by:

Rajma 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 11-NOV-1998 21:35:02.86
 Decayed to: 11-NOV-1998 21:35:02.86
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.55*	5302	1.51	1323.09	1316	16	3.0		CS-137	7.71
0	1172.97	2476	1.79	2345.98	2337	17	4.4		CO-60	4.91
0	1332.18	2144	1.82	2664.44	2655	17	4.4		CO-60	4.72

Summary of Nuclide Activity

Page : 2

Sample ID : WL26735-STD

Acquisition date : 11-NOV-1998 21:35:02

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.815E+00	4.815E+00	0.150E+00	3.11	
CS-137	30.00Y	1.00	7.712E+00	7.712E+00	0.232E+00	3.00	

Total Activity :	1.253E+01	1.253E+01

Grand Total Activity :	1.253E+01	1.253E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26735-STD

Acquisition date : 11-NOV-1998 21:35:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	263.	477.59	6.9846E-01
NA-24	15.	1368.55	4.7123E-02
K-40	101.	1460.75	1.1046E+00
AR-41	23.	1293.60	6.3432E-02
SC-46	115.	1120.55	1.0046E-01
CR-51	287.	320.08	5.1510E-01
MN-54	110.	834.83	7.7589E-02
CO-56	115.	846.76	8.0236E-02
CO-58	109.	810.78	7.5645E-02
FE-59	122.	1099.25	1.8001E-01
ZN-65	101.	1115.55	1.8527E-01
SE-75	336.	264.66	8.1213E-02
KR-85	180.	514.00	1.4853E+01
SR-85	180.	514.01	6.7129E-02
Y-88	3.	1836.06	3.2389E-02
Y-91	41.	1204.67	2.2024E+01
NB-94	124.	871.09	8.5049E-02
NB-95	111.	765.78	7.2594E-02
ZR-95	99.	756.73	1.2484E-01
ZRNB-95	109.	724.18	3.1514E-01
RU-103	213.	497.08	7.4406E-02
RURH-106	133.	621.93	1.3328E+00
AG-108m	110.	722.94	7.5893E-02
CD-109	297.	88.03	1.3887E+00
AG-110M	229.	657.76	9.5297E-02
SN-113	253.	391.69	9.1936E-02
TE-123m	338.	159.00	4.3124E-02
SB-124	138.	602.73	6.7066E-02
SB-125	278.	427.89	2.2753E-01
TE-125m	312.	109.27	1.3874E+01
I-129	270.	39.60	1.5259E+01
I-131	274.	364.48	7.0554E-02
XE-131m	326.	163.93	1.8277E+00
BA-133	255.	356.02	8.6982E-02
CS-134	150.	604.70	7.0075E-02
CS-136	130.	818.51	8.3018E-02
CS-138	10.	1435.86	8.7841E-02
CE-139	342.	165.85	4.5948E-02
BA-140	164.	537.31	2.6297E-01
LA-140	6.	1596.21	3.8552E-02
CE-141	315.	145.44	7.2219E-02
CE-144	334.	133.51	3.2678E-01
CEPR-144	334.	133.51	6.5293E-01
EU-152	300.	121.78	1.2552E-01
EU-154	22.	1274.51	1.4984E-01
EU-155	292.	86.54	1.5597E-01
HF-181	253.	482.18	8.8835E-02
TA-182	281.	67.75	2.1716E-01
HG-203	316.	279.20	5.9177E-02
BI-207	150.	569.70	6.6168E-02
TL-208	351.	277.36	7.9898E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26735-STD

Acquisition date : 11-NOV-1998 21:35:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	282.	46.50	1.0737E+01
BI-212	97.	727.18	9.8183E-01
PB-212	365.	238.63	1.0464E-01
BI-214	152.	609.31	1.5478E-01
PB-214	306.	351.92	2.9169E-01
RA-224	378.	240.99	1.1819E+00
RA-226	387.	186.10	1.1580E+00
AC-228	150.	911.21	3.6308E-01
TH-228	299.	84.37	4.2834E+00
TH-229	302.	88.47	2.0270E-01
U-232	286.	57.78	7.8503E+01
PA-233	286.	312.17	1.3131E-01
UTH-233	371.	245.34	4.2664E+01
PA-234M	146.	1001.03	1.5786E+01
TH-234	281.	63.29	2.9308E+00
U-235	392.	185.71	7.0860E-02
NP-237	292.	86.48	4.1318E-01
U-237	307.	59.54	4.1982E-01
NP-238	144.	984.45	3.6423E-01
NP-239	293.	106.12	1.7106E-01
PU-239	316.	129.30	5.5613E+02
AM-241	307.	59.54	4.0271E-01
AM-243	293.	74.67	1.1666E-01

*****

* 222-S Laboratory Counting Room 12-NOV-1998 01:02:05.08 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: WL26735-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Neuhyt

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2799.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:28.47 sec
 Dead Time: 0.3%

Verified by:

Wjmer 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 11-NOV-1998 22:31:07.36
 Decayed to: 11-NOV-1998 22:31:07.36
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	352.14*	48	1.13	704.28	699	10106.9			
1	509.64	61	1.53	1019.28	1016	13 53.0	7.68E-01		
0	661.43*	32	1.54	1322.85	1317	14139.5		CS-137	1.574E-02
0	1460.54*	31	2.09	2921.20	2911	18140.1		K-40	0.231
0	1764.30*	17	1.54	3528.84	3522	15 88.3			

Summary of Nuclide Activity

Sample ID : WL26735-BLK

Page : 2

Acquisition date : 11-NOV-1998 22:31:07

Total number of lines in spectrum	5	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	5	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~2.315E-01~~	~~2.315E-01~~	~~2.243E-01~~	~~140.07~~			
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~1.574E-02~~	~~1.574E-02~~	~~2.196E-02~~	~~139.55~~			
Total Activity :			2.472E-01	2.472E-01					

Grand Total Activity : 2.472E-01 2.472E-01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : WL26735-BLK

Page : 3

Acquisition date : 11-NOV-1998 22:31:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	96.	477.59	1.4383E-01
NA-24	26.	1368.55	2.0882E-02
AR-41	12.	1293.60	2.1190E-02
SC-46	65.	1120.55	2.5618E-02
CR-51	141.	320.08	1.2199E-01
MN-54	35.	834.83	1.5252E-02
CO-56	41.	846.76	1.6453E-02
CO-58	47.	810.78	1.7011E-02
FE-59	28.	1099.25	3.0240E-02
CO-60	19.	1332.50	1.6724E-02
ZN-65	34.	1115.55	3.7197E-02
SE-75	183.	264.66	2.0205E-02
KR-85	97.	514.00	3.6883E+00
SR-85	97.	514.01	1.6658E-02
Y-88	6.	1836.06	1.4148E-02
Y-91	34.	1204.67	6.7087E+00
NB-94	31.	871.09	1.4906E-02
NB-95	48.	765.78	1.6379E-02
ZR-95	55.	756.73	3.1699E-02
ZRNB-95	42.	724.18	6.7205E-02
RU-103	89.	497.08	1.6340E-02
RURH-106	67.	621.93	3.2097E-01
AG-108m	40.	722.94	1.5831E-02
CD-109	187.	88.03	3.7019E-01
AG-110M	74.	657.76	1.8548E-02
SN-113	119.	391.69	2.1356E-02
TE-123m	228.	159.00	1.1885E-02
SB-124	73.	602.73	1.6479E-02
SB-125	94.	427.89	4.5118E-02
TE-125m	195.	109.27	3.6816E+00
I-129	169.	39.60	4.0660E+00
I-131	96.	364.48	1.4311E-02
XE-131m	231.	163.93	5.1710E-01
BA-133	134.	356.02	2.1341E-02
CS-134	69.	604.70	1.6211E-02
CS-136	42.	818.51	1.6229E-02
CS-138	17.	1435.86	7.5322E-02
CE-139	226.	165.85	1.2547E-02
BA-140	57.	537.31	5.3292E-02
LA-140	11.	1596.21	1.6663E-02
CE-141	219.	145.44	2.0210E-02
CE-144	199.	133.51	8.4942E-02
CEPR-144	199.	133.51	1.6973E-01
EU-152	189.	121.78	3.3450E-02
EU-154	22.	1274.51	5.0132E-02
EU-155	195.	86.54	4.2819E-02
HF-181	81.	482.18	1.7269E-02
TA-182	183.	67.75	5.8880E-02
HG-203	157.	279.20	1.4108E-02
BI-207	61.	569.70	1.4403E-02
TL-208	174.	277.36	1.8972E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26735-BLK

Acquisition date : 11-NOV-1998 22:31:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	144.	46.50	2.5969E+00
BI-212	66.	727.18	2.7171E-01
PB-212	226.	238.63	2.7680E-02
BI-214	181.	609.31	5.5967E-02
PB-214	223.	351.92	1.8539E-01
RA-224	208.	240.99	2.9704E-01
RA-226	258.	186.10	3.1700E-01
AC-228	71.	911.21	8.4586E-02
TH-228	197.	84.37	1.1692E+00
TH-229	186.	88.47	5.3571E-02
U-232	176.	57.78	2.0716E+01
PA-233	137.	312.17	3.0751E-02
UTH-233	177.	245.34	9.9624E+00
PA-234M	43.	1001.03	2.9529E+00
TH-234	184.	63.29	7.9577E-01
U-235	249.	185.71	1.8956E-02
NP-237	195.	86.48	1.1341E-01
U-237	174.	59.54	1.0671E-01
NP-238	46.	984.45	7.1827E-02
NP-239	173.	106.12	4.4672E-02
PU-239	199.	129.30	1.4841E+02
AM-241	173.	59.54	1.0199E-01
AM-243	217.	74.67	3.3647E-02

*****
 * 222-S Laboratory Counting Room 12-NOV-1998 19:44:46.04
 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T002814-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2806.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:06.77 sec
 Dead Time: 6.3%

Verified by:

11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 12-NOV-1998 17:04:11.73
 Decayed to: 12-NOV-1998 17:04:11.73
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	74.72*	901	1.16	149.50	148	6 81.5		AM-243	77.1
0	661.55*	2072080	1.49	1323.09	1315	16 0.1		CS-137	1.625E+05
0	1322.84	210	2.02	2645.76	2640	13 14.6			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002814-SAM

Acquisition date : 12-NOV-1998 17:04:11

Total number of lines in spectrum	3
Number of unidentified lines	1
Number of lines tentatively identified by NID	2 66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.625E+05	1.625E+05	0.002E+05	0.14	
~~AM-243~~	~~7380.00Y*****~~		~~7.709E+01~~	~~7.709E+01~~	~~6.283E+01~~	~~81.50~~	~~11-16-98~~
Total Activity :			1.626E+05	1.626E+05			

Grand Total Activity : 1.626E+05   1.626E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002814-SAM

Acquisition date : 12-NOV-1998 17:04:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	62189.	477.59	5.6526E+02
NA-24	19.	1368.55	2.8513E+00
K-40	320.	1460.75	9.9778E+01
AR-41	26.	1293.60	4.7433E+00
SC-46	89.	1120.55	4.7169E+00
CR-51	58170.	320.08	3.9441E+02
MN-54	118.	834.83	4.2951E+00
CO-56	103.	846.76	4.0778E+00
CO-58	126.	810.78	4.3543E+00
FE-59	49.	1099.25	6.2254E+00
CO-60	27.	1332.50	3.0960E+00
ZN-65	45.	1115.55	6.7297E+00
SE-75	65570.	264.66	6.1969E+01
KR-85	32226.	514.00	1.0362E+04
SR-85	32220.	514.01	4.6871E+01
Y-88	11.	1836.06	2.7764E+00
Y-91	37.	1204.67	1.1080E+03
NB-94	80.	871.09	3.7075E+00
NB-95	222.	765.78	5.4417E+00
ZR-95	255.	756.73	1.0549E+01
ZRNB-95	466.	724.18	3.4078E+01
RU-103	39792.	497.08	5.3342E+01
RURH-106	14077.	621.93	7.0884E+02
AG-108m	487.	722.94	8.3406E+00
CD-109	54992.	88.03	1.1102E+03
AG-110M	45713.	657.76	7.0083E+01
SN-113	59194.	391.69	7.4635E+01
TE-123m	66705.	159.00	3.4535E+01
SB-124	14433.	602.73	3.5453E+01
SB-125	67672.	427.89	1.8794E+02
TE-125m	59659.	109.27	1.1216E+04
I-129	49938.	39.60	1.0670E+04
I-131	56825.	364.48	5.4317E+01
XE-131m	66663.	163.93	1.4883E+03
BA-133	56927.	356.02	6.9359E+01
CS-134	14376.	604.70	3.5535E+01
CS-136	99.	818.51	3.9072E+00
CS-138	20.	1435.86	1.3460E+01
CE-139	67121.	165.85	3.6590E+01
BA-140	23690.	537.31	1.6500E+02
LA-140	7.	1596.21	2.1909E+00
CE-141	68239.	145.44	6.1067E+01
CE-144	67687.	133.51	2.6915E+02
CEPR-144	67688.	133.51	5.3782E+02
EU-152	65167.	121.78	1.0746E+02
EU-154	29.	1274.51	8.9923E+00
EU-155	54656.	86.54	1.2545E+02
HF-181	50843.	482.18	6.6305E+01
TA-182	50388.	67.75	1.6911E+02
HG-203	63355.	279.20	4.5570E+01
BI-207	15748.	569.70	3.5240E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002814-SAM

Acquisition date : 12-NOV-1998 17:04:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	63667.	277.36	5.8629E+02
PB-210	49128.	46.50	7.7226E+03
BI-212	474.	727.18	1.1293E+02
PB-212	70248.	238.63	8.0065E+01
BI-214	14063.	609.31	7.7068E+01
PB-214	56703.	351.92	4.9980E+02
RA-224	69899.	240.99	8.9215E+02
RA-226	85729.	186.10	9.7070E+02
AC-228	104.	911.21	1.6302E+01
TH-228	54392.	84.37	3.3957E+03
TH-229	55144.	88.47	1.6112E+02
U-232	49416.	57.78	5.8768E+04
PA-233	58670.	312.17	1.0129E+02
UTH-233	69096.	245.34	3.2055E+04
PA-234M	57.	1001.03	5.3953E+02
TH-234	49915.	63.29	2.2520E+03
U-235	84550.	185.71	5.8638E+01
NP-237	54658.	86.48	3.3229E+02
U-237	49331.	59.54	3.0629E+02
NP-238	73.	984.45	1.4267E+01
NP-239	58647.	106.12	1.4312E+02
PU-239	67262.	129.30	4.6966E+05
AM-241	49330.	59.54	2.9265E+02

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 04:01:31.65
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T2814-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2809.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:01.78 sec
 Dead Time: 6.3%

Verified by:

11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 01:21:01.14
 Decayed to: 13-NOV-1998 01:21:01.14
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.55*	2062132	1.48	1323.09	1315	16	0.1		CS-137	1.617E+05
0	1322.81	192	1.66	2645.70	2639	16	21.2			
0	1460.59*	29	1.78	2921.31	2911	17	154.7		K-40	33.1

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2814-DUP

Acquisition date : 13-NOV-1998 01:21:01

Total number of lines in spectrum	3
Number of unidentified lines	1
Number of lines tentatively identified by NID	2 66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~3.311E+01~~	~~3.311E+01~~	~~5.122E+01~~	~~154.69~~	~~MS~~
CS-137	30.00Y	1.00	1.617E+05	1.617E+05	0.002E+05	0.14	
Total Activity :			1.618E+05	1.618E+05			11-16-98

Grand Total Activity : 1.618E+05   1.618E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2814-DUP

Acquisition date : 13-NOV-1998 01:21:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	61933.	477.59	5.6409E+02
NA-24	21.	1368.55	3.0233E+00
AR-41	21.	1293.60	4.3815E+00
SC-46	76.	1120.55	4.3697E+00
CR-51	57723.	320.08	3.9289E+02
MN-54	99.	834.83	3.9647E+00
CO-56	115.	846.76	4.2995E+00
CO-58	101.	810.78	3.9273E+00
FE-59	46.	1099.25	6.0689E+00
CO-60	27.	1332.50	3.0952E+00
ZN-65	66.	1115.55	8.0591E+00
SE-75	65323.	264.66	6.1852E+01
KR-85	32178.	514.00	1.0355E+04
SR-85	32177.	514.01	4.6840E+01
Y-88	5.	1836.06	1.9729E+00
Y-91	45.	1204.67	1.2063E+03
NB-94	94.	871.09	3.9868E+00
NB-95	222.	765.78	5.4389E+00
ZR-95	268.	756.73	1.0804E+01
ZRNB-95	461.	724.18	3.3917E+01
RU-103	39357.	497.08	5.3050E+01
RURH-106	13839.	621.93	7.0286E+02
AG-108m	474.	722.94	8.2360E+00
CD-109	54627.	88.03	1.1065E+03
AG-110M	45301.	657.76	6.9767E+01
SN-113	58958.	391.69	7.4487E+01
TE-123m	66831.	159.00	3.4568E+01
SB-124	14024.	602.73	3.4950E+01
SB-125	66677.	427.89	1.8656E+02
TE-125m	59682.	109.27	1.1218E+04
I-129	49597.	39.60	1.0634E+04
I-131	56848.	364.48	5.4328E+01
XE-131m	66753.	163.93	1.4893E+03
BA-133	56324.	356.02	6.8992E+01
CS-134	14127.	604.70	3.5227E+01
CS-136	111.	818.51	4.1284E+00
CS-138	11.	1435.86	1.0143E+01
CE-139	66927.	165.85	3.6538E+01
BA-140	23694.	537.31	1.6502E+02
LA-140	19.	1596.21	3.2920E+00
CE-141	67806.	145.44	6.0873E+01
CE-144	67366.	133.51	2.6851E+02
CEPR-144	67364.	133.51	5.3654E+02
EU-152	65397.	121.78	1.0765E+02
EU-154	21.	1274.51	7.7138E+00
EU-155	54657.	86.54	1.2546E+02
HF-181	50403.	482.18	6.6019E+01
TA-182	50048.	67.75	1.6854E+02
HC-203	63127.	279.20	4.5488E+01
BI-207	15494.	569.70	3.4955E+01
TL-208	63249.	277.36	5.8436E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2814-DUP

Acquisition date : 13-NOV-1998 01:21:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	49468.	46.50	7.7492E+03
BI-212	445.	727.18	1.0951E+02
PB-212	70095.	238.63	7.9978E+01
BI-214	14030.	609.31	7.6977E+01
PB-214	56347.	351.92	4.9799E+02
RA-224	69549.	240.99	8.8992E+02
RA-226	84902.	186.10	9.6602E+02
AC-228	97.	911.21	1.5790E+01
TH-228	54333.	84.37	3.3938E+03
TH-229	54703.	88.47	1.6047E+02
U-232	49053.	57.78	5.8552E+04
PA-233	58522.	312.17	1.0117E+02
UTH-233	68772.	245.34	3.1980E+04
PA-234M	54.	1001.03	5.2651E+02
TH-234	49058.	63.29	2.2327E+03
U-235	83745.	185.71	5.8358E+01
NP-237	54663.	86.48	3.3231E+02
U-237	49134.	59.54	3.0568E+02
NP-238	58.	984.45	1.2845E+01
NP-239	58631.	106.12	1.4309E+02
PU-239	67064.	129.30	4.6897E+05
AM-241	49133.	59.54	2.9207E+02
AM-243	52208.	74.67	9.1189E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 06:52:53.74 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T002815-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2810.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:55.94 sec
 Dead Time: 6.8%

Verified by:

PJ Jones 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 04:11:28.48
 Decayed to: 13-NOV-1998 04:11:28.48
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	188.44	4646	4.39	376.92	372	12 35.3			
0	661.54*	2255968	1.49	1323.08	1315	16 0.1		CS-137	1.769E+05
0	1322.59	244	1.94	2645.27	2636	19 16.0			
0	1391.06	23	1.86	2782.21	2775	12 52.2			
0	1764.73*	19	2.60	3529.69	3522	17 80.3			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002815-SAM

Acquisition date : 13-NOV-1998 04:11:28

Total number of lines in spectrum	5	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	2	40.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.769E+05	1.769E+05	0.002E+05	0.14	
Total Activity :			1.769E+05	1.769E+05			

Grand Total Activity :	1.769E+05	1.769E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002815-SAM

Acquisition date : 13-NOV-1998 04:11:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	68423.	477.59	5.9285E+02
NA-24	19.	1368.55	2.8779E+00
K-40	311.	1460.75	9.8409E+01
AR-41	22.	1293.60	4.3940E+00
SC-46	71.	1120.55	4.2355E+00
CR-51	63754.	320.08	4.1286E+02
MN-54	118.	834.83	4.2935E+00
CO-56	108.	846.76	4.1740E+00
CO-58	127.	810.78	4.3670E+00
FE-59	58.	1099.25	6.7214E+00
CO-60	34.	1332.50	3.4301E+00
ZN-65	51.	1115.55	7.1187E+00
SE-75	71425.	264.66	6.4670E+01
KR-85	35146.	514.00	1.0820E+04
SR-85	35140.	514.01	4.8942E+01
Y-88	7.	1836.06	2.2842E+00
Y-91	42.	1204.67	1.1758E+03
NB-94	96.	871.09	4.0416E+00
NB-95	266.	765.78	5.9356E+00
ZR-95	301.	756.73	1.1419E+01
ZRNB-95	532.	724.18	3.6360E+01
RU-103	42548.	497.08	5.5152E+01
RURH-106	15208.	621.93	7.3663E+02
AG-108m	537.	722.94	8.7514E+00
CD-109	59504.	88.03	1.1547E+03
AG-110M	50286.	657.76	7.3495E+01
SN-113	64392.	391.69	7.7835E+01
TE-123m	73353.	159.00	3.6212E+01
SB-124	15545.	602.73	3.6787E+01
SB-125	73179.	427.89	1.9542E+02
TE-125m	65033.	109.27	1.1709E+04
I-129	54698.	39.60	1.1166E+04
I-131	62540.	364.48	5.6978E+01
XE-131m	73438.	163.93	1.5620E+03
BA-133	62181.	356.02	7.2482E+01
CS-134	15423.	604.70	3.6800E+01
CS-136	110.	818.51	4.1172E+00
CS-138	20.	1435.86	1.3455E+01
CE-139	73338.	165.85	3.8244E+01
BA-140	25974.	537.31	1.7275E+02
LA-140	11.	1596.21	2.5567E+00
CE-141	73952.	145.44	6.3567E+01
CE-144	73813.	133.51	2.8104E+02
CEPR-144	73813.	133.51	5.6157E+02
EU-152	71503.	121.78	1.1255E+02
EU-154	30.	1274.51	9.1797E+00
EU-155	59363.	86.54	1.3073E+02
HF-181	55486.	482.18	6.9260E+01
TA-182	54583.	67.75	1.7599E+02
HG-203	68856.	279.20	4.7502E+01
BI-207	16877.	569.70	3.6476E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002815-SAM

Acquisition date : 13-NOV-1998 04:11:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	69562.	277.36	6.1277E+02
PB-210	53986.	46.50	8.0947E+03
BI-212	504.	727.18	1.1637E+02
PB-212	76756.	238.63	8.3684E+01
BI-214	15654.	609.31	8.1290E+01
PB-214	62145.	351.92	5.2568E+02
RA-224	76207.	240.99	9.3150E+02
RA-226	93447.	186.10	1.0134E+03
AC-228	133.	911.21	1.8314E+01
TH-228	59187.	84.37	3.5418E+03
TH-229	59561.	88.47	1.6743E+02
U-232	53645.	57.78	6.1224E+04
PA-233	64261.	312.17	1.0600E+02
UTH-233	74778.	245.34	3.3344E+04
PA-234M	49.	1001.03	5.0606E+02
TH-234	54330.	63.29	2.3492E+03
U-235	92225.	185.71	6.1236E+01
NP-237	59368.	86.48	3.4628E+02
U-237	53749.	59.54	3.1968E+02
NP-238	72.	984.45	1.4188E+01
NP-239	64028.	106.12	1.4953E+02
PU-239	73365.	129.30	4.9046E+05
AM-241	53749.	59.54	3.0544E+02
AM-243	56997.	74.67	9.5268E+01

*****

* 222-S Laboratory Counting Room 13-NOV-1998 09:58:39.88 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26735
 Sample ID: S98T2815-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Miller

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2811.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:56.99 sec
 Dead Time: 6.8%

Verified by:

R. Jones 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 13-NOV-1998 07:17:14.16
 Decayed to: 13-NOV-1998 07:17:14.16
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	189.02	9262	5.86	378.07	371	17 23.9			
0	661.53*	2252346	1.49	1323.05	1315	16 0.1		CS-137	1.766E+05
0	1322.79	223	1.76	2645.65	2638	18 16.1			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2815-DUP

Acquisition date : 13-NOV-1998 07:17:14

Total number of lines in spectrum	3	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	1	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.766E+05	1.766E+05	0.002E+05	0.14	
Total Activity :			1.766E+05	1.766E+05			

Grand Total Activity :	1.766E+05	1.766E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2815-DUP

Acquisition date : 13-NOV-1998 07:17:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	67635.	477.59	5.8943E+02
NA-24	11.	1368.55	2.3070E+00
K-40	309.	1460.75	9.7990E+01
AR-41	28.	1293.60	4.9170E+00
SC-46	68.	1120.55	4.1666E+00
CR-51	63841.	320.08	4.1315E+02
MN-54	110.	834.83	4.1592E+00
CO-56	106.	846.76	4.1406E+00
CO-58	145.	810.78	4.6669E+00
FE-59	64.	1099.25	7.0249E+00
CO-60	32.	1332.50	3.3307E+00
ZN-65	49.	1115.55	7.0278E+00
SE-75	71521.	264.66	6.4713E+01
KR-85	35424.	514.00	1.0863E+04
SR-85	35421.	514.01	4.9136E+01
Y-88	5.	1836.06	2.0074E+00
Y-91	36.	1204.67	1.0965E+03
NB-94	93.	871.09	3.9789E+00
NB-95	257.	765.78	5.8438E+00
ZR-95	305.	756.73	1.1502E+01
ZRNB-95	554.	724.18	3.7077E+01
RU-103	43247.	497.08	5.5603E+01
RURH-106	15052.	621.93	7.3288E+02
AG-108m	575.	722.94	9.0475E+00
CD-109	59964.	88.03	1.1591E+03
AG-110M	52268.	657.76	7.4926E+01
SN-113	64693.	391.69	7.8017E+01
TE-123m	73551.	159.00	3.6260E+01
SB-124	15369.	602.73	3.6580E+01
SB-125	72872.	427.89	1.9501E+02
TE-125m	65266.	109.27	1.1730E+04
I-129	54332.	39.60	1.1129E+04
I-131	62278.	364.48	5.6858E+01
XE-131m	73044.	163.93	1.5578E+03
BA-133	62061.	356.02	7.2412E+01
CS-134	15506.	604.70	3.6899E+01
CS-136	123.	818.51	4.3343E+00
CS-138	18.	1435.86	1.2742E+01
CE-139	73443.	165.85	3.8271E+01
BA-140	25874.	537.31	1.7242E+02
LA-140	13.	1596.21	2.7388E+00
CE-141	73874.	145.44	6.3533E+01
CE-144	73894.	133.51	2.8119E+02
CEPR-144	73893.	133.51	5.6188E+02
EU-152	71674.	121.78	1.1268E+02
EU-154	30.	1274.51	9.1584E+00
EU-155	59598.	86.54	1.3099E+02
HF-181	55167.	482.18	6.9061E+01
TA-182	54440.	67.75	1.7576E+02
HG-203	68964.	279.20	4.7540E+01
BI-207	17088.	569.70	3.6701E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2815-DUP

Acquisition date : 13-NOV-1998 07:17:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	69218.	277.36	6.1125E+02
PB-210	53593.	46.50	8.0652E+03
BI-212	519.	727.18	1.1807E+02
PB-212	76959.	238.63	8.3794E+01
BI-214	15583.	609.31	8.1107E+01
PB-214	62474.	351.92	5.2712E+02
RA-224	76002.	240.99	9.3026E+02
RA-226	93442.	186.10	1.0133E+03
AC-228	96.	911.21	1.5651E+01
TH-228	59381.	84.37	3.5476E+03
TH-229	60017.	88.47	1.6807E+02
U-232	53653.	57.78	6.1229E+04
PA-233	64089.	312.17	1.0586E+02
UTH-233	75640.	245.34	3.3536E+04
PA-234M	65.	1001.03	5.7651E+02
TH-234	54034.	63.29	2.3429E+03
U-235	92166.	185.71	6.1217E+01
NP-237	59603.	86.48	3.4696E+02
U-237	53424.	59.54	3.1872E+02
NP-238	72.	984.45	1.4158E+01
NP-239	63842.	106.12	1.4932E+02
PU-239	73439.	129.30	4.9070E+05
AM-241	53424.	59.54	3.0452E+02
AM-243	56643.	74.67	9.4973E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 12:45:02.18 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T2816-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SL Label

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2812.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:42:09.55 sec
 Dead Time: 7.5%

Verified by:

JD Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 10:01:58.81
 Decayed to: 13-NOV-1998 10:01:58.81
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	84.80	950	1.31	169.66	168	5	74.8		TH-228	2.966E+03
0	187.93	3208	3.30	375.91	372	10	46.2			
0	661.52*	2519239	1.49	1323.04	1315	16	0.1		CS-137	1.976E+05
0	1322.76	341	1.78	2645.61	2637	18	13.6			
0	1460.57*	33	2.09	2921.26	2914	15	128.1		K-40	37.8

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2816-SAM

Acquisition date : 13-NOV-1998 10:01:58

Total number of lines in spectrum	5	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	3	60.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~3.782E+01~~	~~3.782E+01~~	~~4.845E+01~~	~~128.09~~	~~KAG~~
CS-137	30.00Y	1.00	1.976E+05	1.976E+05	0.003E+05	0.13	
~~TH-232~~	~~1.91Y~~	~~1.00~~	~~2.966E+03~~	~~2.966E+03~~	~~2.218E+03~~	~~74.80~~	~~KAG~~
Total Activity :			2.006E+05	2.006E+05			

11-16-98

Grand Total Activity :	2.006E+05	2.006E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2816-SAM

Acquisition date : 13-NOV-1998 10:01:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	76253.	477.59	6.2578E+02
NA-24	18.	1368.55	2.8239E+00
AR-41	25.	1293.60	4.7252E+00
SC-46	97.	1120.55	4.9061E+00
CR-51	70551.	320.08	4.3427E+02
MN-54	132.	834.83	4.5369E+00
CO-56	147.	846.76	4.8272E+00
CO-58	148.	810.78	4.7029E+00
FE-59	56.	1099.25	6.6164E+00
CO-60	34.	1332.50	3.4208E+00
ZN-65	75.	1115.55	8.5373E+00
SE-75	79462.	264.66	6.8204E+01
KR-85	39208.	514.00	1.1427E+04
SR-85	39204.	514.01	5.1686E+01
Y-88	6.	1836.06	2.1320E+00
Y-91	44.	1204.67	1.1989E+03
NB-94	125.	871.09	4.5691E+00
NB-95	309.	765.78	6.3859E+00
ZR-95	366.	756.73	1.2566E+01
ZRNB-95	654.	724.18	4.0221E+01
RU-103	48288.	497.08	5.8746E+01
RURH-106	16984.	621.93	7.7827E+02
AG-108m	677.	722.94	9.7952E+00
CD-109	66635.	88.03	1.2218E+03
AG-110M	58470.	657.76	7.9236E+01
SN-113	71951.	391.69	8.2267E+01
TE-123m	82440.	159.00	3.8385E+01
SB-124	17474.	602.73	3.8993E+01
SB-125	81917.	427.89	2.0674E+02
TE-125m	73298.	109.27	1.2429E+04
I-129	60498.	39.60	1.1742E+04
I-131	69276.	364.48	5.9963E+01
XE-131m	81873.	163.93	1.6491E+03
BA-133	68458.	356.02	7.6044E+01
CS-134	17293.	604.70	3.8957E+01
CS-136	132.	818.51	4.4964E+00
CS-138	14.	1435.86	1.1596E+01
CE-139	81846.	165.85	4.0397E+01
BA-140	28580.	537.31	1.8118E+02
LA-140	11.	1596.21	2.6014E+00
CE-141	82172.	145.44	6.7000E+01
CE-144	82358.	133.51	2.9683E+02
CEPR-144	82357.	133.51	5.9312E+02
EU-152	80086.	121.78	1.1910E+02
EU-154	26.	1274.51	8.6188E+00
EU-155	66452.	86.54	1.3830E+02
HF-181	62175.	482.18	7.3307E+01
TA-182	61135.	67.75	1.8623E+02
HG-203	77272.	279.20	5.0316E+01
BI-207	18960.	569.70	3.8651E+01
TL-208	77556.	277.36	6.4695E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2816-SAM

Acquisition date : 13-NOV-1998 10:01:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	60037.	46.50	8.5355E+03
BI-212	627.	727.18	1.2950E+02
PB-212	86053.	238.63	8.8597E+01
BI-214	17193.	609.31	8.5175E+01
PB-214	69039.	351.92	5.5795E+02
RA-224	85404.	240.99	9.8608E+02
RA-226	104551.	186.10	1.0718E+03
AC-228	125.	911.21	1.7756E+01
TH-229	66721.	88.47	1.7719E+02
U-232	59608.	57.78	6.4529E+04
PA-233	71795.	312.17	1.1203E+02
UTH-233	84112.	245.34	3.5360E+04
PA-234M	76.	1001.03	6.1919E+02
TH-234	60734.	63.29	2.4835E+03
U-235	103328.	185.71	6.4811E+01
NP-237	66483.	86.48	3.6640E+02
U-237	59740.	59.54	3.3700E+02
NP-238	72.	984.45	1.4158E+01
NP-239	71253.	106.12	1.5775E+02
PU-239	81574.	129.30	5.1711E+05
AM-241	59741.	59.54	3.2198E+02
AM-243	63465.	74.67	1.0052E+02

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 15:40:49.48 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T2816-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2814.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:42:11.57 sec
 Dead Time: 7.5%

Verified by:

11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 12:58:08.81
 Decayed to: 13-NOV-1998 12:58:08.81
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	189.35	17003	6.39	378.73	370	24	18.7			
0	661.52*	2523891	1.49	1323.03	1315	16	0.1		CS-137	1.979E+05
0	1322.96	300	2.09	2646.01	2639	17	15.8			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2816-DUP

Acquisition date : 13-NOV-1998 12:58:08

Total number of lines in spectrum	3	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	1	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.979E+05	1.979E+05	0.003E+05	0.13	
Total Activity :			1.979E+05	1.979E+05			

Grand Total Activity :	1.979E+05	1.979E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2816-DUP

Acquisition date : 13-NOV-1998 12:58:08

Nuclide	Bkgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	76295.	477.59	6.2595E+02
NA-24	14.	1368.55	2.5399E+00
K-40	332.	1460.75	1.0154E+02
AR-41	23.	1293.60	4.5155E+00
SC-46	93.	1120.55	4.8201E+00
CR-51	70993.	320.08	4.3563E+02
MN-54	124.	834.83	4.4044E+00
CO-56	126.	846.76	4.4857E+00
CO-58	156.	810.78	4.8305E+00
FE-59	77.	1099.25	7.6808E+00
CO-60	28.	1332.50	3.1507E+00
ZN-65	67.	1115.55	8.1311E+00
SE-75	80357.	264.66	6.8586E+01
KR-85	39480.	514.00	1.1466E+04
SR-85	39472.	514.01	5.1862E+01
Y-88	8.	1836.06	2.3985E+00
Y-91	42.	1204.67	1.1707E+03
NB-94	105.	871.09	4.2113E+00
NB-95	291.	765.78	6.1997E+00
ZR-95	390.	756.73	1.2958E+01
ZRNB-95	658.	724.18	4.0324E+01
RU-103	48503.	497.08	5.8876E+01
RURH-106	16861.	621.93	7.7545E+02
AG-108m	674.	722.94	9.7790E+00
CD-109	66722.	88.03	1.2226E+03
AG-110M	59322.	657.76	7.9810E+01
SN-113	72560.	391.69	8.2614E+01
TE-123m	82553.	159.00	3.8411E+01
SB-124	17421.	602.73	3.8934E+01
SB-125	82030.	427.89	2.0688E+02
TE-125m	73398.	109.27	1.2437E+04
I-129	60991.	39.60	1.1789E+04
I-131	69513.	364.48	6.0065E+01
XE-131m	82302.	163.93	1.6534E+03
BA-133	69663.	356.02	7.6709E+01
CS-134	17355.	604.70	3.9026E+01
CS-136	143.	818.51	4.6624E+00
CS-138	15.	1435.86	1.1991E+01
CE-139	82256.	165.85	4.0497E+01
BA-140	28732.	537.31	1.8166E+02
LA-140	12.	1596.21	2.6728E+00
CE-141	83118.	145.44	6.7384E+01
CE-144	82360.	133.51	2.9683E+02
CEPR-144	82362.	133.51	5.9314E+02
EU-152	79320.	121.78	1.1853E+02
EU-154	37.	1274.51	1.0066E+01
EU-155	66489.	86.54	1.3834E+02
HF-181	62123.	482.18	7.3276E+01
TA-182	61552.	67.75	1.8686E+02
HG-203	76895.	279.20	5.0193E+01
BI-207	19432.	569.70	3.9127E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2816-DUP

Acquisition date : 13-NOV-1998 12:58:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	76973.	277.36	6.4451E+02
PB-210	59746.	46.50	8.5148E+03
BI-212	619.	727.18	1.2863E+02
PB-212	85899.	238.63	8.8518E+01
BI-214	17059.	609.31	8.4844E+01
PB-214	69233.	351.92	5.5884E+02
RA-224	85586.	240.99	9.8713E+02
RA-226	104608.	186.10	1.0721E+03
AC-228	126.	911.21	1.7835E+01
TH-228	66345.	84.37	3.7494E+03
TH-229	66875.	88.47	1.7739E+02
U-232	60286.	57.78	6.4894E+04
PA-233	71687.	312.17	1.1195E+02
UTH-233	84358.	245.34	3.5412E+04
PA-234M	72.	1001.03	6.0204E+02
TH-234	60303.	63.29	2.4747E+03
U-235	103291.	185.71	6.4799E+01
NP-237	66515.	86.48	3.6648E+02
U-237	60604.	59.54	3.3943E+02
NP-238	55.	984.45	1.2483E+01
NP-239	71889.	106.12	1.5845E+02
PU-239	82273.	129.30	5.1931E+05
AM-241	60604.	59.54	3.2429E+02
AM-243	63427.	74.67	1.0049E+02

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 18:37:18.35 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T2817-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Newlight

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2815.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:41:43.86 sec
 Dead Time: 7.3%

Verified by:

RJmer 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 15:55:05.14
 Decayed to: 13-NOV-1998 15:55:05.14
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	27.45	8232	1.58	54.96	54	5	8.5			
0	661.51*	2426159	1.49	1323.02	1315	16	0.1		CS-137	1.903E+05
0	1322.66	294	2.00	2645.41	2639	13	13.5			
0	1629.66	10	0.93	3259.50	3254	8	93.8			
0	1764.66*	17	1.96	3529.56	3523	15	88.3			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2817-SAM

Acquisition date : 13-NOV-1998 15:55:05

Total number of lines in spectrum	5	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	2	40.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.903E+05	1.903E+05	0.003E+05	0.13	
Total Activity :			1.903E+05	1.903E+05			

Grand Total Activity :	1.903E+05	1.903E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2817-SAM

Acquisition date : 13-NOV-1998 15:55:05

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	73345.	477.59	6.1376E+02
NA-24	26.	1368.55	3.2877E+00
K-40	327.	1460.75	1.0076E+02
AR-41	29.	1293.60	5.0182E+00
SC-46	83.	1120.55	4.5517E+00
CR-51	68459.	320.08	4.2779E+02
MN-54	131.	834.83	4.5250E+00
CO-56	106.	846.76	4.1318E+00
CO-58	142.	810.78	4.6100E+00
FE-59	45.	1099.25	6.0082E+00
CO-60	37.	1332.50	3.5650E+00
ZN-65	65.	1115.55	7.9959E+00
SE-75	76929.	264.66	6.7110E+01
KR-85	37705.	514.00	1.1206E+04
SR-85	37700.	514.01	5.0688E+01
Y-88	5.	1836.06	1.9729E+00
Y-91	47.	1204.67	1.2378E+03
NB-94	112.	871.09	4.3379E+00
NB-95	307.	765.78	6.3690E+00
ZR-95	358.	756.73	1.2438E+01
ZRNB-95	584.	724.18	3.8044E+01
RU-103	46580.	497.08	5.7700E+01
RURH-106	16586.	621.93	7.6913E+02
AG-108m	612.	722.94	9.3228E+00
CD-109	64550.	88.03	1.2025E+03
AG-110M	57566.	657.76	7.8622E+01
SN-113	69277.	391.69	8.0728E+01
TE-123m	78943.	159.00	3.7563E+01
SB-124	16818.	602.73	3.8257E+01
SB-125	78263.	427.89	2.0208E+02
TE-125m	70242.	109.27	1.2168E+04
I-129	58246.	39.60	1.1521E+04
I-131	67096.	364.48	5.9013E+01
XE-131m	79059.	163.93	1.6205E+03
BA-133	66589.	356.02	7.5001E+01
CS-134	16750.	604.70	3.8343E+01
CS-136	142.	818.51	4.6427E+00
CS-138	8.	1435.86	9.2486E+00
CE-139	79108.	165.85	3.9717E+01
BA-140	27628.	537.31	1.7815E+02
LA-140	7.	1596.21	2.1039E+00
CE-141	79628.	145.44	6.5957E+01
CE-144	79643.	133.51	2.9190E+02
CEPR-144	79640.	133.51	5.8327E+02
EU-152	76534.	121.78	1.1643E+02
EU-154	25.	1274.51	8.4179E+00
EU-155	64633.	86.54	1.3640E+02
HF-181	59351.	482.18	7.1626E+01
TA-182	58765.	67.75	1.8259E+02
HG-203	74014.	279.20	4.9246E+01
BI-207	18503.	569.70	3.8184E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2817-SAM

Acquisition date : 13-NOV-1998 15:55:05

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	74390.	277.36	6.3363E+02
PB-210	57844.	46.50	8.3784E+03
BI-212	574.	727.18	1.2401E+02
PB-212	82141.	238.63	8.6564E+01
BI-214	16760.	609.31	8.4102E+01
PB-214	66572.	351.92	5.4656E+02
RA-224	81420.	240.99	9.6282E+02
RA-226	100232.	186.10	1.0494E+03
AC-228	142.	911.21	1.8877E+01
TH-228	64535.	84.37	3.6980E+03
TH-229	64552.	88.47	1.7429E+02
U-232	57688.	57.78	6.3484E+04
PA-233	69064.	312.17	1.0988E+02
UTH-233	81129.	245.34	3.4729E+04
PA-234M	72.	1001.03	6.0428E+02
TH-234	58271.	63.29	2.4328E+03
U-235	99052.	185.71	6.3458E+01
NP-237	64649.	86.48	3.6132E+02
U-237	57671.	59.54	3.3112E+02
NP-238	70.	984.45	1.4002E+01
NP-239	69220.	106.12	1.5548E+02
PU-239	78715.	129.30	5.0799E+05
AM-241	57671.	59.54	3.1636E+02
AM-243	61197.	74.67	9.8708E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 21:39:19.38 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26735
 Sample ID: S98T002817-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

NEWIGHT

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2816.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:41:47.06 sec
 Dead Time: 7.3%

Verified by:

W Jones 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 18:57:03.29
 Decayed to: 13-NOV-1998 18:57:03.29
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	34.09	942	1.18	68.24	66	6 87.8			
0	188.27	7330	4.94	376.58	371	15 28.1			
0	661.52*	2439621	1.49	1323.03	1315	16 0.1		CS-137	1.913E+05
0	1322.87	280	1.78	2645.82	2638	14 13.8			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002817-DUP

Acquisition date : 13-NOV-1998 18:57:03

Total number of lines in spectrum	4
Number of unidentified lines	3
Number of lines tentatively identified by NID	1 25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.913E+05	1.913E+05	0.003E+05	0.13	
Total Activity :			1.913E+05	1.913E+05			

Grand Total Activity :	1.913E+05	1.913E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002817-DUP

Acquisition date : 13-NOV-1998 18:57:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	73377.	477.59	6.1389E+02
NA-24	22.	1368.55	3.0589E+00
K-40	296.	1460.75	9.6006E+01
AR-41	34.	1293.60	5.4340E+00
SC-46	96.	1120.55	4.8865E+00
CR-51	68479.	320.08	4.2786E+02
MN-54	124.	834.83	4.4019E+00
CO-56	126.	846.76	4.4854E+00
CO-58	139.	810.78	4.5694E+00
FE-59	63.	1099.25	7.0190E+00
CO-60	31.	1332.50	3.2957E+00
ZN-65	69.	1115.55	8.1987E+00
SE-75	77177.	264.66	6.7218E+01
KR-85	38224.	514.00	1.1283E+04
SR-85	38216.	514.01	5.1032E+01
Y-88	7.	1836.06	2.2625E+00
Y-91	47.	1204.67	1.2328E+03
NB-94	112.	871.09	4.3311E+00
NB-95	291.	765.78	6.2070E+00
ZR-95	356.	756.73	1.2401E+01
ZRNB-95	643.	724.18	3.9874E+01
RU-103	46319.	497.08	5.7538E+01
RURH-106	16130.	621.93	7.5854E+02
AG-108m	659.	722.94	9.6717E+00
CD-109	64430.	88.03	1.2014E+03
AG-110M	57315.	657.76	7.8452E+01
SN-113	69812.	391.69	8.1038E+01
TE-123m	79296.	159.00	3.7647E+01
SB-124	16720.	602.73	3.8147E+01
SB-125	79341.	427.89	2.0347E+02
TE-125m	70274.	109.27	1.2171E+04
I-129	58850.	39.60	1.1581E+04
I-131	67285.	364.48	5.9096E+01
XE-131m	80216.	163.93	1.6323E+03
BA-133	66746.	356.02	7.5089E+01
CS-134	16775.	604.70	3.8371E+01
CS-136	139.	818.51	4.6080E+00
CS-138	24.	1435.86	1.4630E+01
CE-139	79224.	165.85	3.9745E+01
BA-140	27781.	537.31	1.7864E+02
LA-140	11.	1596.21	2.6141E+00
CE-141	79491.	145.44	6.5900E+01
CE-144	79583.	133.51	2.9179E+02
CEPR-144	79583.	133.51	5.8307E+02
EU-152	77020.	121.78	1.1680E+02
EU-154	34.	1274.51	9.7128E+00
EU-155	64464.	86.54	1.3622E+02
HF-181	60051.	482.18	7.2046E+01
TA-182	59141.	67.75	1.8317E+02
HG-203	74547.	279.20	4.9422E+01
BI-207	18318.	569.70	3.7994E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002817-DUP

Acquisition date : 13-NOV-1998 18:57:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	74587.	277.36	6.3446E+02
PB-210	57887.	46.50	8.3815E+03
BI-212	627.	727.18	1.2951E+02
PB-212	82957.	238.63	8.6992E+01
BI-214	16709.	609.31	8.3973E+01
PB-214	66770.	351.92	5.4754E+02
RA-224	82611.	240.99	9.6983E+02
RA-226	101137.	186.10	1.0542E+03
AC-228	123.	911.21	1.7671E+01
TH-228	64204.	84.37	3.6886E+03
TH-229	64524.	88.47	1.7425E+02
U-232	57737.	57.78	6.3511E+04
PA-233	69680.	312.17	1.1037E+02
UTH-233	81695.	245.34	3.4849E+04
PA-234M	53.	1001.03	5.2236E+02
TH-234	58759.	63.29	2.4429E+03
U-235	99916.	185.71	6.3734E+01
NP-237	64450.	86.48	3.6076E+02
U-237	58212.	59.54	3.3267E+02
NP-238	72.	984.45	1.4202E+01
NP-239	69262.	106.12	1.5553E+02
PU-239	78823.	129.30	5.0833E+05
AM-241	58212.	59.54	3.1784E+02
AM-243	61536.	74.67	9.8981E+01

LABCORE Completed Worklist Report for Worklist# 26737

Analyst: gll

Instrument: GEA01

Book#: 28B57

Method: LA-548-121

Rev/Mod

F-φ

Worklist Comment: BY-105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	GEA-01 C060-02	LIQUID	4.529e-03	4.68e-03	103.334	% Recovery
1 STD	0	0	GEA-01 C060-02E	LIQUID	1	3.16	3.160	% Ct Error
1 STD	0	0	GEA-01 CS13702	LIQUID	7.453e-03	7.43e-03	99.691	% Recovery
1 STD	0	0	GEA-01 CS13702E	LIQUID	1	2.84	2.840	% Ct Error
2 BLNK	0	0	GEA-01 C060-02	LIQUID	1	<1.24e-3		uCi/mL
2 BLNK	0	0	GEA-01 CS13702	LIQUID	1	<1.52e-3		uCi/mL
3 SAMPLE	S98T002818	0	GEA-01 C060-02	LIQUID	N/A	4.041e-03	4.04e-003	uCi/mL
3 SAMPLE	S98T002818	0	GEA-01 C060-02E	LIQUID	N/A	n/a		% Ct. Error
3 SAMPLE	S98T002818	0	GEA-01 CS13702	LIQUID	N/A	2.180e+02		uCi/mL
3 SAMPLE	S98T002818	0	GEA-01 CS13702E	LIQUID	N/A	0.110		% Ct. Error
4 DUP	S98T002818	0	GEA-01 C060-02	LIQUID	<4.04e-3	<3.77e-3		RPD
4 DUP	S98T002818	0	GEA-01 C060-02E	LIQUID	1	n/a		% Ct Error
4 DUP	S98T002818	0	GEA-01 CS13702	LIQUID	2.18e+02	2.20e+02	0.913	RPD
4 DUP	S98T002818	0	GEA-01 CS13702E	LIQUID	1	0.110	0.110	% Ct Error
5 SAMPLE	S98T002819	0	GEA-01 C060-02	LIQUID	N/A	3.846e-03	3.85e-003	uCi/mL
5 SAMPLE	S98T002819	0	GEA-01 C060-02E	LIQUID	N/A	n/a		% Ct. Error
5 SAMPLE	S98T002819	0	GEA-01 CS13702	LIQUID	N/A	2.170e+02		uCi/mL
5 SAMPLE	S98T002819	0	GEA-01 CS13702E	LIQUID	N/A	0.110		% Ct. Error
6 DUP	S98T002819	0	GEA-01 C060-02	LIQUID	<3.85e-3	<3.87e-3		RPD
6 DUP	S98T002819	0	GEA-01 C060-02E	LIQUID	1	n/a		% Ct Error
6 DUP	S98T002819	0	GEA-01 CS13702	LIQUID	2.17e+02	2.20e+02	1.373	RPD
6 DUP	S98T002819	0	GEA-01 CS13702E	LIQUID	1	0.110	0.110	% Ct Error
7 SAMPLE	S98T002820	0	GEA-01 C060-02	LIQUID	N/A	1.346e-02	1.35e-002	uCi/mL
7 SAMPLE	S98T002820	0	GEA-01 C060-02E	LIQUID	N/A	n/a		% Ct. Error
7 SAMPLE	S98T002820	0	GEA-01 CS13702	LIQUID	N/A	2.530e+02		uCi/mL
7 SAMPLE	S98T002820	0	GEA-01 CS13702E	LIQUID	N/A	0.190		% Ct. Error
8 DUP	S98T002820	0	GEA-01 C060-02	LIQUID	<1.35e-2	<1.34e-2		RPD
8 DUP	S98T002820	0	GEA-01 C060-02E	LIQUID	1	n/a		% Ct Error
8 DUP	S98T002820	0	GEA-01 CS13702	LIQUID	2.53e+02	2.51e+02	0.794	RPD
8 DUP	S98T002820	0	GEA-01 CS13702E	LIQUID	1	0.190	0.190	% Ct Error
9 SAMPLE	S98T002821	0	GEA-01 C060-02	LIQUID	N/A	1.147e-02	1.15e-002	uCi/mL
9 SAMPLE	S98T002821	0	GEA-01 C060-02E	LIQUID	N/A	n/a		% Ct. Error
9 SAMPLE	S98T002821	0	GEA-01 CS13702	LIQUID	N/A	2.940e+02		uCi/mL
9 SAMPLE	S98T002821	0	GEA-01 CS13702E	LIQUID	N/A	0.180		% Ct. Error
10 DUP	S98T002821	0	GEA-01 C060-02	LIQUID	<1.15e-2	<1.32e-2		RPD
10 DUP	S98T002821	0	GEA-01 C060-02E	LIQUID	1	n/a		% Ct Error
10 DUP	S98T002821	0	GEA-01 CS13702	LIQUID	2.94e+02	2.97e+02	1.015	RPD
10 DUP	S98T002821	0	GEA-01 CS13702E	LIQUID	1	0.180	0.180	% Ct Error
11 SAMPLE	S98T002939	0	GEA-01 C060-02	LIQUID	N/A	4.039e-03	4.04e-003	uCi/mL
11 SAMPLE	S98T002939	0	GEA-01 C060-02E	LIQUID	N/A	n/a		% Ct. Error
11 SAMPLE	S98T002939	0	GEA-01 CS13702	LIQUID	N/A	2.330e+02		uCi/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26737

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	SAMPLE	S98T002939	0		@GEA-01	CS13702E LIQUID	N/A	0.110		% Ct. Error
12	DUP	S98T002939	0		@GEA-01	CO60-02 LIQUID	<4.04e-3	<4.23e-3		RPD
12	DUP	S98T002939	0		@GEA-01	CO60-02E LIQUID	1	n/a		% Ct Error
12	DUP	S98T002939	0		@GEA-01	CS13702 LIQUID	2.33e+02	2.27e+02	2.609	RPD
12	DUP	S98T002939	0		@GEA-01	CS13702E LIQUID	1	0.110	0.110	% Ct Error

Comments Section:

Comments for sample# S98T002818 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002819 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002820 and test @GEA-01 .
DL=0 => n/a

Comments for sample# S98T002821 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002939 and test @GEA-01 .
DL=0 => n/a.

Final page for worklist# 26737

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11-16-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26737

Analyst: gll Instrument: GEA00 1 Book# 28B57

Method: LA-548-121 Rev/Mod F-8

Worklist Comment: BY-105 FOR @GEA-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	LIQUID		
2 BLNK			@GEA-01	LIQUID		
3 SAMPLE	S98T002818 0		@GEA-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4 DUP	S98T002818 0		@GEA-01	LIQUID		
5 SAMPLE	S98T002819 0		@GEA-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6 DUP	S98T002819 0		@GEA-01	LIQUID		
7 SAMPLE	S98T002820 0		@GEA-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8 DUP	S98T002820 0		@GEA-01	LIQUID		
9 SAMPLE	S98T002821 0		@GEA-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
10 DUP	S98T002821 0		@GEA-01	LIQUID		
11 SAMPLE	S98T002939 0		@GEA-01	LIQUID	98000490	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
12 DUP	S98T002939 0		@GEA-01	LIQUID		

Final page for worklist # 26737

Gay L. Platt 11-12-98
Analyst Signature Date

[Signature] 11/16/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****

* 222-S Laboratory Counting Room 13-NOV-1998 02:22:53.99 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: WL26737-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Reulight

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]lg1054.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:15.16 sec
 Dead Time: 0.5%

Verified by:

Ph Jones 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 01:32:07.10
 Decayed to: 13-NOV-1998 01:32:07.10
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.85	95	1.40	25.82	24	5 65.1			
0	31.97	1085	1.05	64.06	60	9 11.9			
0	36.41	304	0.88	72.94	69	9 35.8			
0	59.57	239	1.04	119.25	116	8 43.5		AM-241	0.166
0	661.72*	6132	1.41	1323.54	1316	14 2.8		CS-137	7.42
0	1173.21	2807	1.76	2346.71	2339	17 4.3		CO-60	4.83
0	1332.45	2373	1.98	2665.30	2655	21 4.6		CO-60	4.53

Summary of Nuclide Activity

Page : 2

Sample ID : WL26737-STD

Acquisition date : 13-NOV-1998 01:32:07

Total number of lines in spectrum	7	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	4	57.14%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.680E+00	4.680E+00	0.148E+00	3.16	
CS-137	30.00Y	1.00	7.425E+00	7.425E+00	0.211E+00	2.84	
AM-241	432.20Y*****		1.657E-01	1.657E-01	0.720E-01	43.46	
Total Activity :			1.227E+01	1.227E+01			

Grand Total Activity :	1.227E+01	1.227E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26737-STD

Acquisition date : 13-NOV-1998 01:32:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	380.	477.59	6.8722E-01
NA-24	19.	1368.55	4.6035E-02
K-40	602.	1460.75	2.2425E+00
AR-41	49.	1293.60	7.7032E-02
SC-46	185.	1120.55	1.0917E-01
CR-51	438.	320.08	5.2893E-01
MN-54	178.	834.83	8.2738E-02
CO-56	202.	846.76	8.9013E-02
CO-58	192.	810.78	8.3941E-02
FE-59	164.	1099.25	1.7960E-01
ZN-65	148.	1115.55	1.9280E-01
SE-75	497.	264.66	8.2669E-02
KR-85	278.	514.00	1.5083E+01
SR-85	278.	514.01	6.8160E-02
Y-88	3.	1836.06	2.6148E-02
Y-91	69.	1204.67	2.4151E+01
NB-94	213.	871.09	9.3822E-02
NB-95	161.	765.78	7.2961E-02
ZR-95	145.	756.73	1.2577E-01
ZRNB-95	182.	724.18	3.3699E-01
RU-103	297.	497.08	7.1998E-02
RURH-106	208.	621.93	1.3711E+00
AG-108m	178.	722.94	7.9915E-02
CD-109	644.	88.03	8.4103E-01
AG-110M	338.	657.76	9.5634E-02
SN-113	383.	391.69	9.3115E-02
TE-123m	663.	159.00	4.6399E-02
SB-124	209.	602.73	6.7715E-02
SB-125	444.	427.89	2.3584E-01
TE-125m	694.	109.27	1.1848E+01
I-129	439.	39.60	3.8293E-01
I-131	424.	364.48	7.2422E-02
XE-131m	683.	163.93	2.0559E+00
BA-133	408.	356.02	9.0918E-02
CS-134	202.	604.70	6.6944E-02
CS-136	198.	818.51	8.5783E-02
CS-138	19.	1435.86	9.8653E-02
CE-139	658.	165.85	4.9848E-02
BA-140	231.	537.31	2.5571E-01
LA-140	18.	1596.21	5.2362E-02
CE-141	695.	145.44	7.8746E-02
CE-144	747.	133.51	3.3962E-01
CEPR-144	748.	133.51	6.7870E-01
EU-152	670.	121.78	1.2057E-01
EU-154	39.	1274.51	1.7028E-01
EU-155	637.	86.54	9.1765E-02
HF-181	352.	482.18	8.5955E-02
TA-182	533.	67.75	6.5713E-02
HG-203	512.	279.20	6.2965E-02
BI-207	229.	569.70	6.7102E-02
TL-208	527.	277.36	8.1968E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26737-STD

Acquisition date : 13-NOV-1998 01:32:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	500.	46.50	7.0094E-01
BI-212	207.	727.18	1.1795E+00
PB-212	628.	238.63	1.1464E-01
BI-214	341.	609.31	1.8895E-01
PB-214	549.	351.92	3.2213E-01
RA-224	606.	240.99	1.2515E+00
RA-226	764.	186.10	1.3159E+00
AC-228	291.	911.21	4.2593E-01
TH-228	676.	84.37	2.4376E+00
TH-229	631.	88.47	1.2186E-01
U-232	623.	57.78	1.4844E+01
PA-233	461.	312.17	1.3858E-01
UTH-233	589.	245.34	4.5007E+01
PA-234M	216.	1001.03	1.6409E+01
TH-234	541.	63.29	7.2084E-01
U-235	759.	185.71	7.9775E-02
NP-237	642.	86.48	2.4376E-01
U-237	817.	59.54	9.7993E-02
NP-238	195.	984.45	3.6229E-01
NP-239	648.	106.12	1.4043E-01
PU-239	767.	129.30	5.8602E+02
AM-243	655.	74.67	5.0003E-02

*****

* 222-S Laboratory Counting Room 13-NOV-1998 05:05:13.65 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26737
 Sample ID: WL26737-BLK
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Shulab

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1055.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:30.89 sec
 Dead Time: 0.3%

Verified by:

H Jones 11-16-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 13-NOV-1998 02:34:13.59
 Decayed to: 13-NOV-1998 02:34:13.59
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.74	145	0.78	25.61	24	5 63.0			
0	16.56	460	1.70	33.24	29	12 38.5			
0	294.60*	140	1.13	589.27	584	10 83.9			

Summary of Nuclide Activity

Page : 2

Sample ID : WL26737-BLK

Acquisition date : 13-NOV-1998 02:34:13

Total number of lines in spectrum	3
Number of unidentified lines	2
Number of lines tentatively identified by NID	1
	33.33%

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26737-BLK

Acquisition date : 13-NOV-1998 02:34:13

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	308.	477.59	9.1100E+00
NA-24	55.	1368.55	1.1235E+00
K-40	1702.	1460.75	5.4765E+01
AR-41	106.	1293.60	2.1616E+00
SC-46	283.	1120.55	1.9662E+00
CR-51	576.	320.08	8.8679E+00
MN-54	174.	834.83	1.2005E+00
CO-56	167.	846.76	1.1944E+00
CO-58	169.	810.78	1.1607E+00
FE-59	144.	1099.25	2.4714E+00
CO-60	80.	1332.50	1.2360E+00
ZN-65	164.	1115.55	2.9692E+00
SE-75	815.	264.66	1.5448E+00
KR-85	342.	514.00	2.4434E+02
SR-85	341.	514.01	1.1038E+00
Y-88	15.	1836.06	7.3533E-01
Y-91	141.	1204.67	4.9834E+02
NB-94	133.	871.09	1.0974E+00
NB-95	201.	765.78	1.1931E+00
ZR-95	177.	756.73	2.0283E+00
ZRNB-95	195.	724.18	5.1021E+00
RU-103	290.	497.08	1.0439E+00
RURH-106	227.	621.93	2.0965E+01
AG-108m	185.	722.94	1.1939E+00
CD-109	1244.	88.03	1.7043E+01
AG-110M	223.	657.76	1.1480E+00
SN-113	421.	391.69	1.4305E+00
TE-123m	1371.	159.00	9.7229E-01
SB-124	211.	602.73	9.9795E-01
SB-125	398.	427.89	3.2820E+00
TE-125m	1360.	109.27	2.4177E+02
I-129	814.	39.60	7.5927E+00
I-131	464.	364.48	1.1136E+00
XE-131m	1304.	163.93	4.1501E+01
BA-133	489.	356.02	1.4566E+00
CS-134	232.	604.70	1.0481E+00
CS-136	145.	818.51	1.0851E+00
CS-137	318.	661.66	1.5208E+00
CS-138	45.	1435.86	4.4281E+00
CE-139	1281.	165.85	1.0136E+00
BA-140	295.	537.31	4.2355E+00
LA-140	36.	1596.21	1.0502E+00
CE-141	1371.	145.44	1.6139E+00
CE-144	1481.	133.51	6.9702E+00
CEPR-144	1481.	133.51	1.3928E+01
EU-152	1382.	121.78	2.5231E+00
EU-154	82.	1274.51	3.5075E+00
EU-155	1251.	86.54	1.8748E+00
HF-181	321.	482.18	1.2052E+00
TA-182	917.	67.75	1.2577E+00
HG-203	735.	279.20	1.1018E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26737-BLK

Acquisition date : 13-NOV-1998 02:34:13

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	268.	569.70	1.0619E+00
TL-208	792.	277.36	1.4669E+01
PB-210	852.	46.50	1.3378E+01
BI-212	234.	727.18	1.8326E+01
PB-212	1210.	238.63	2.3203E+00
BI-214	643.	609.31	3.7723E+00
PB-214	829.	351.92	1.2857E+01
RA-224	986.	240.99	2.3446E+01
RA-226	1295.	186.10	2.5011E+01
AC-228	376.	911.21	7.0801E+00
TH-228	1324.	84.37	4.9734E+01
TH-229	1238.	88.47	2.4864E+00
U-232	1021.	57.78	2.7732E+02
PA-233	666.	312.17	2.4336E+00
UTH-233	828.	245.34	7.7958E+02
PA-234M	154.	1001.03	2.0491E+02
TH-234	1071.	63.29	1.4766E+01
U-235	1308.	185.71	1.5282E+00
NP-237	1256.	86.48	4.9682E+00
U-237	1021.	59.54	1.6087E+00
NP-238	138.	984.45	4.5462E+00
NP-239	1322.	106.12	2.9514E+00
PU-239	1520.	129.30	1.2028E+04
AM-241	1021.	59.54	1.5377E+00
AM-243	1346.	74.67	1.0439E+00

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 09:07:31.30 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2818-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1056.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:45:36.28 sec
 Dead Time: 9.4%

Verified by:

R. Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 06:21:26.58
 Decayed to: 13-NOV-1998 06:21:26.58
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	9.42	4808	1.53	18.97	18	5 13.3			
0	12.47	1679	0.58	25.07	23	6 46.8			
0	21.82	18257	1.20	43.75	40	8 6.0			
0	31.96	525863	1.13	64.03	60	9 0.4			
0	36.46	146013	1.12	73.03	69	9 1.0			
2	72.96*	3217	1.12	146.03	143	11 29.1	2.90E+00		
2	74.98*	5695	0.94	150.06	143	11 14.7	Pb x-rays	AM-243	156.
0	84.59*	3564	1.22	169.28	167	7 26.5		TH-228	4.609E+03
0	190.79	94842	10.32	381.68	368	33 4.8			
0	661.68*	3412506	1.53	1323.45	1315	17 0.1		CS-137	2.185E+05
0	692.49	920	2.81	1385.10	1379	16 25.8			
0	698.30	103	1.40	1396.70	1394	10100.0			
0	1323.02	536	2.19	2646.43	2638	15 13.1			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2818-SAM

Acquisition date : 13-NOV-1998 06:21:26

Total number of lines in spectrum 13
 Number of unidentified lines 10
 Number of lines tentatively identified by NID 3 23.08%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.185E+05	2.185E+05	0.002E+05	0.11	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~4.609E+03~~	~~4.609E+03~~	~~1.221E+03~~	~~26.48~~	~~raj~~
~~AM-243~~	~~7380.00Y*****~~	~~1.558E+02~~	~~1.558E+02~~	~~1.558E+02~~	~~0.228E+02~~	~~14.66~~	~~raj~~
Total Activity :			2.232E+05	2.232E+05			11-16-98

Grand Total Activity : 2.232E+05 2.232E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2818-SAM

Acquisition date : 13-NOV-1998 06:21:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	106479.	477.59	6.0790E+02
NA-24	60.	1368.55	3.7001E+00
K-40	1685.	1460.75	1.7175E+02
AR-41	116.	1293.60	7.4029E+00
SC-46	286.	1120.55	6.3869E+00
CR-51	82972.	320.08	3.8205E+02
MN-54	265.	834.83	5.0705E+00
CO-56	250.	846.76	4.9873E+00
CO-58	256.	810.78	4.9175E+00
FE-59	184.	1099.25	9.0224E+00
CO-60	86.	1332.50	4.0415E+00
ZN-65	190.	1115.55	1.0321E+01
SE-75	80312.	264.66	5.4952E+01
KR-85	53397.	514.00	1.0963E+04
SR-85	53392.	514.01	4.9588E+01
Y-88	21.	1836.06	2.9000E+00
Y-91	140.	1204.67	1.5823E+03
NB-94	199.	871.09	4.5521E+00
NB-95	283.	765.78	4.9423E+00
ZR-95	261.	756.73	8.6213E+00
ZRNB-95	297.	724.18	2.2213E+01
RU-103	64776.	497.08	5.5942E+01
RURH-106	16800.	621.93	6.3383E+02
AG-108m	300.	722.94	5.3540E+00
CD-109	64716.	88.03	5.3701E+02
AG-110M	108277.	657.76	8.7995E+01
SN-113	95438.	391.69	7.7548E+01
TE-123m	79784.	159.00	2.7814E+01
SB-124	17340.	602.73	3.1847E+01
SB-125	112980.	427.89	1.9923E+02
TE-125m	70006.	109.27	7.1561E+03
I-129	55928.	39.60	1.6972E+02
I-131	88572.	364.48	5.5347E+01
XE-131m	80991.	163.93	1.2196E+03
BA-133	87205.	356.02	6.9983E+01
CS-134	17311.	604.70	3.1952E+01
CS-136	230.	818.51	4.7048E+00
CS-138	35.	1435.86	1.3523E+01
CE-139	81553.	165.85	3.0085E+01
BA-140	37932.	537.31	1.7154E+02
LA-140	41.	1596.21	3.5837E+00
CE-141	76608.	145.44	4.6094E+01
CE-144	75137.	133.51	1.9378E+02
CEPR-144	75138.	133.51	3.8721E+02
EU-152	74421.	121.78	7.4051E+01
EU-154	84.	1274.51	1.1287E+01
EU-155	65757.	86.54	5.9605E+01
HF-181	84879.	482.18	7.0415E+01
TA-182	60337.	67.75	4.5507E+01
HG-203	81001.	279.20	4.1446E+01
BI-207	21917.	569.70	3.4118E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2818-SAM

Acquisition date : 13-NOV-1998 06:21:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	80734.	277.36	5.3080E+02
PB-210	56380.	46.50	3.8653E+02
BI-212	337.	727.18	7.7658E+01
PB-212	82119.	238.63	6.8744E+01
BI-214	17302.	609.31	7.0024E+01
PB-214	86549.	351.92	5.1927E+02
RA-224	81425.	240.99	7.6566E+02
RA-226	131754.	186.10	9.2253E+02
AC-228	423.	911.21	2.5360E+01
TH-229	64355.	88.47	7.8248E+01
U-232	57779.	57.78	8.9223E+03
PA-233	82533.	312.17	9.7309E+01
UTH-233	80765.	245.34	2.7594E+04
PA-234M	186.	1001.03	7.4280E+02
TH-234	59151.	63.29	4.8478E+02
U-235	129088.	185.71	5.5522E+01
NP-237	65939.	86.48	1.5788E+02
U-237	58359.	59.54	5.2718E+01
NP-238	188.	984.45	1.7603E+01
NP-239	67121.	106.12	8.7628E+01
PU-239	75263.	129.30	3.3326E+05
AM-241	58360.	59.54	5.0364E+01

*****

* 222-S Laboratory Counting Room 13-NOV-1998 11:59:12.78 *

*****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2818-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

J. L. Hader

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1057.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:45:47.64 sec
 Dead Time: 9.5%

Verified by:

R. Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 09:12:56.97
 Decayed to: 13-NOV-1998 09:12:56.97
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
2	8.93	13218	0.89	17.99	16	7	2.4	4.46E+02		
2	9.83	6334	0.89	19.78	16	7	8.7			
0	12.35	2589	0.61	24.83	22	7	34.3			
0	21.82	17887	1.30	43.77	40	8	6.2			
0	31.95	533311	1.13	64.02	60	9	0.4			
0	36.45	147415	1.13	73.02	69	9	1.0			
1	72.90*	2681	1.01	145.91	143	11	31.0	2.54E+00		
1	74.98*	5901	1.02	150.07	143	11	14.2		AM-243	161.
0	84.76*	2548	1.07	169.62	168	6	33.1		TH-228	3.295E+03
0	87.34*	1050	0.81	174.79	173	7	89.4		NP-237	139.
									CD-109	476.
0	189.78	14569	5.36	379.64	373	15	16.7			
2	651.77	2178	1.82	1303.63	1297	37	23.7	1.98E+03		
2	661.69*	3435779	1.59	1323.48	1297	37	0.1		CS-137	2.200E+05
0	692.37	1037	3.02	1384.84	1378	14	20.7			
0	1323.10	592	2.03	2646.59	2636	21	12.2			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2818-DUP

Acquisition date : 13-NOV-1998 09:12:56

Total number of lines in spectrum 15
 Number of unidentified lines 11
 Number of lines tentatively identified by NID 4 26.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CD-109~~	~~462.90D~~	~~1.00~~	~~4.757E+02~~	~~4.758E+02~~	~~4.253E+02~~	~~89.39~~	~~RAJ~~
CS-137	30.00Y	1.00	2.200E+05	2.200E+05	0.002E+05	0.11	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~3.295E+03~~	~~3.295E+03~~	~~1.092E+03~~	~~33.15~~	~~RAJ~~
~~HP-237~~	~~2.14E+06Y~~	~~1.00~~	~~1.386E+02~~	~~1.386E+02~~	~~1.239E+02~~	~~89.39~~	~~RAJ~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~1.615E+02~~	~~1.615E+02~~	~~0.230E+02~~	~~14.23~~	~~RAJ~~
Total Activity :			2.240E+05	2.240E+05			

11-16-98

Grand Total Activity : 2.240E+05 2.240E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : S98T2818-DUP

Page : 3

Acquisition date : 13-NOV-1998 09:12:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	108091.	477.59	6.1248E+02
NA-24	64.	1368.55	3.8364E+00
K-40	1720.	1460.75	1.7350E+02
AR-41	88.	1293.60	6.4912E+00
SC-46	310.	1120.55	6.6385E+00
CR-51	84354.	320.08	3.8521E+02
MN-54	207.	834.83	4.5022E+00
CO-56	260.	846.76	5.0791E+00
CO-58	201.	810.78	4.3799E+00
FE-59	171.	1099.25	8.7235E+00
CO-60	74.	1332.50	3.7740E+00
ZN-65	179.	1115.55	1.0008E+01
SE-75	81419.	264.66	5.5329E+01
KR-85	54059.	514.00	1.1030E+04
SR-85	54051.	514.01	4.9892E+01
Y-88	26.	1836.06	3.1753E+00
Y-91	154.	1204.67	1.6602E+03
NB-94	221.	871.09	4.7910E+00
NB-95	270.	765.78	4.8307E+00
ZR-95	281.	756.73	8.9330E+00
ZRNB-95	283.	724.18	2.1711E+01
RU-103	66100.	497.08	5.6510E+01
RURH-106	16882.	621.93	6.3538E+02
AG-108m	288.	722.94	5.2427E+00
AG-110M	110064.	657.76	8.8717E+01
SN-113	97321.	391.69	7.8307E+01
TE-123m	80885.	159.00	2.8005E+01
SB-124	17544.	602.73	3.2033E+01
SB-125	114174.	427.89	2.0028E+02
TE-125m	70554.	109.27	7.1840E+03
I-129	56841.	39.60	1.7110E+02
I-131	89646.	364.48	5.5681E+01
XE-131m	82034.	163.93	1.2275E+03
BA-133	88208.	356.02	7.0384E+01
CS-134	17484.	604.70	3.2111E+01
CS-136	213.	818.51	4.5271E+00
CS-138	41.	1435.86	1.4519E+01
CE-139	82980.	165.85	3.0346E+01
BA-140	38216.	537.31	1.7218E+02
LA-140	34.	1596.21	3.2738E+00
CE-141	77918.	145.44	4.6485E+01
CE-144	76177.	133.51	1.9511E+02
CEPR-144	76174.	133.51	3.8986E+02
EU-152	75579.	121.78	7.4624E+01
EU-154	100.	1274.51	1.2223E+01
EU-155	67065.	86.54	6.0193E+01
HF-181	85911.	482.18	7.0841E+01
TA-182	61474.	67.75	4.5933E+01
HG-203	81558.	279.20	4.1589E+01
BI-207	21751.	569.70	3.3989E+01
TL-208	81729.	277.36	5.3406E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2818-DUP

Acquisition date : 13-NOV-1998 09:12:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	56925.	46.50	3.8839E+02
BI-212	340.	727.18	7.7988E+01
PB-212	82662.	238.63	6.8970E+01
BI-214	17693.	609.31	7.0807E+01
PB-214	87750.	351.92	5.2342E+02
RA-224	81870.	240.99	7.6776E+02
RA-226	132926.	186.10	9.2662E+02
AC-228	445.	911.21	2.6003E+01
TH-229	65796.	88.47	7.9117E+01
U-232	58410.	57.78	8.9708E+03
PA-233	83457.	312.17	9.7851E+01
UTH-233	81411.	245.34	2.7704E+04
PA-234M	220.	1001.03	8.0437E+02
TH-234	60060.	63.29	4.8848E+02
U-235	130127.	185.71	5.5744E+01
U-237	58996.	59.54	5.3004E+01
NP-238	168.	984.45	1.6662E+01
NP-239	67575.	106.12	8.7925E+01
PU-239	76010.	129.30	3.3491E+05
AM-241	58996.	59.54	5.0637E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 15:38:24.31 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2819-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]lg1059.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:45:28.10 sec
 Dead Time: 9.3%

Verified by:

RAJmer 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 12:52:26.59
 Decayed to: 13-NOV-1998 12:52:26.59
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	9.33	4168	1.26	18.79	18	5 15.3			
0	12.36	1584	0.63	24.85	23	5 42.9			
0	21.82	17530	1.22	43.77	40	8 6.3			
0	31.94	520051	1.13	64.00	60	9 0.4			
0	36.45	143772	1.12	73.02	69	9 1.0			
1	72.89*	2935	1.01	145.89	143	11 28.1	6.45E-01		
1	74.98*	5365	1.02	150.07	143	11 15.5		AM-243	147.
0	84.67*	3030	1.00	169.45	168	6 27.5		TH-228	3.919E+03
0	87.42*	1394	0.82	174.96	173	7 66.4		NP-237	184.
								CD-109	632.
0	190.22	49813	8.24	380.53	370	25 7.4			
4	473.55	2463	1.78	947.18	945	8 33.4	9.52E-01		
4	475.14	1006	1.13	950.36	945	8 80.8			
0	661.67*	3389936	1.53	1323.43	1315	17 0.1		CS-137	2.170E+05
0	693.03	592	2.43	1386.16	1382	12 31.3			
0	1322.91	505	1.71	2646.21	2638	16 12.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2819-SAM

Acquisition date : 13-NOV-1998 12:52:26

Total number of lines in spectrum 15
 Number of unidentified lines 11
 Number of lines tentatively identified by NID 4 26.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~GD-109~~	~~462.90D~~	~~1.00~~	~~6.315E+02~~	~~6.316E+02~~	~~4.192E+02~~	~~66.38~~	~~raj~~
CS-137	30.00Y	1.00	2.170E+05	2.170E+05	0.002E+05	0.11	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~3.919E+03~~	~~3.919E+03~~	~~1.077E+03~~	~~27.49~~	~~raj~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~1.840E+02~~	~~1.840E+02~~	~~1.221E+02~~	~~66.38~~	~~raj~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~1.468E+02~~	~~1.468E+02~~	~~0.228E+02~~	~~15.52~~	~~raj~~
Total Activity :			2.219E+05	2.219E+05			11-16-98

Grand Total Activity : 2.219E+05 2.219E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
 Sample ID : S98T2819-SAM

Page : 3
 Acquisition date : 13-NOV-1998 12:52:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	105101.	477.59	6.0396E+02
NA-24	56.	1368.55	3.6040E+00
K-40	1737.	1460.75	1.7434E+02
AR-41	78.	1293.60	6.1456E+00
SC-46	322.	1120.55	6.7644E+00
CR-51	82643.	320.08	3.8129E+02
MN-54	219.	834.83	4.6284E+00
CO-56	253.	846.76	5.0170E+00
CO-58	218.	810.78	4.5503E+00
FE-59	174.	1099.25	8.7916E+00
CO-60	77.	1332.50	3.8460E+00
ZN-65	180.	1115.55	1.0045E+01
SE-75	79948.	264.66	5.4827E+01
KR-85	52645.	514.00	1.0885E+04
SR-85	52638.	514.01	4.9238E+01
Y-88	23.	1836.06	2.9852E+00
Y-91	149.	1204.67	1.6318E+03
NB-94	216.	871.09	4.7379E+00
NB-95	270.	765.78	4.8324E+00
ZR-95	277.	756.73	8.8827E+00
ZRNB-95	309.	724.18	2.2661E+01
RU-103	64891.	497.08	5.5992E+01
RURH-106	16241.	621.93	6.2325E+02
AG-108m	313.	722.94	5.4585E+00
AG-110M	109269.	657.76	8.8397E+01
SN-113	95162.	391.69	7.7436E+01
TE-123m	79069.	159.00	2.7690E+01
SB-124	17153.	602.73	3.1676E+01
SB-125	111677.	427.89	1.9808E+02
TE-125m	69571.	109.27	7.1339E+03
I-129	56187.	39.60	1.7011E+02
I-131	88484.	364.48	5.5319E+01
XE-131m	80884.	163.93	1.2188E+03
BA-133	86364.	356.02	6.9645E+01
CS-134	16961.	604.70	3.1629E+01
CS-136	246.	818.51	4.8598E+00
CS-138	45.	1435.86	1.5161E+01
CE-139	81002.	165.85	2.9983E+01
BA-140	36809.	537.31	1.6899E+02
LA-140	39.	1596.21	3.4842E+00
CE-141	75756.	145.44	4.5837E+01
CE-144	74877.	133.51	1.9345E+02
CEPR-144	74879.	133.51	3.8654E+02
EU-152	74350.	121.78	7.4015E+01
EU-154	114.	1274.51	1.2994E+01
EU-155	65503.	86.54	5.9490E+01
HF-181	83743.	482.18	6.9943E+01
TA-182	59983.	67.75	4.5373E+01
HG-203	80298.	279.20	4.1267E+01
BI-207	21651.	569.70	3.3911E+01
TL-208	80130.	277.36	5.2882E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2819-SAM

Acquisition date : 13-NOV-1998 12:52:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	56156.	46.50	3.8576E+02
BI-212	347.	727.18	7.8789E+01
PB-212	81737.	238.63	6.8584E+01
BI-214	17408.	609.31	7.0237E+01
PB-214	86265.	351.92	5.1802E+02
RA-224	80773.	240.99	7.6259E+02
RA-226	130424.	186.10	9.1787E+02
AC-228	415.	911.21	2.5131E+01
TH-229	64103.	88.47	7.8095E+01
U-232	57678.	57.78	8.9146E+03
PA-233	81902.	312.17	9.6937E+01
UTH-233	80500.	245.34	2.7549E+04
PA-234M	200.	1001.03	7.6921E+02
TH-234	58988.	63.29	4.8411E+02
U-235	128005.	185.71	5.5289E+01
U-237	57879.	59.54	5.2501E+01
NP-238	175.	984.45	1.7007E+01
NP-239	66960.	106.12	8.7522E+01
PU-239	74997.	129.30	3.3267E+05
AM-241	57879.	59.54	5.0157E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 18:40:26.01 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2819-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Neu Light

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1060.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:45:38.68 sec
 Dead Time: 9.4%

Verified by:

RJmer 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 15:54:17.60
 Decayed to: 13-NOV-1998 15:54:17.60
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	9.34	4421	1.36	18.80	18	5 14.5			
0	12.39	1572	0.58	24.90	23	6 50.2			
0	21.79	17149	1.19	43.71	40	8 6.5			
0	31.94	525559	1.14	64.01	60	9 0.4			
0	36.45	145529	1.12	73.02	69	9 1.0			
3	72.84*	2946	1.23	145.79	143	11 24.5	1.60E+00		
3	74.97*	5236	0.90	150.04	143	11 13.9		AM-243	143.
0	84.55*	2245	1.04	169.21	168	6 37.3		TH-228	2.903E+03
0	87.28*	1347	1.00	174.67	173	6 61.0		NP-237	178.
								CD-109	610.
0	190.53	70728	8.86	381.15	369	29 5.9			
0	661.66*	3429515	1.53	1323.42	1315	17 0.1		CS-137	2.196E+05
0	692.82	701	2.48	1385.74	1381	13 28.1			
0	1323.09	528	2.12	2646.56	2640	15 11.0			
0	1764.28*	31	3.03	3529.31	3520	17 96.2	1.70E+00		

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2819-DUP

Acquisition date : 13-NOV-1998 15:54:17

Total number of lines in spectrum 14
 Number of unidentified lines 9
 Number of lines tentatively identified by NID 5 35.71%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~GD-109~~	~~462.90D~~	~~1.00~~	~~6.099E+02~~	~~6.100E+02~~	~~3.720E+02~~	~~60.99~~	~~RAJ~~
CS-137	30.00Y	1.00	2.196E+05	2.196E+05	0.002E+05	0.11	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~2.903E+03~~	~~2.903E+03~~	~~1.083E+03~~	~~37.31~~	~~RAJ~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~1.777E+02~~	~~1.777E+02~~	~~1.084E+02~~	~~60.99~~	~~RAJ~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~1.433E+02~~	~~1.433E+02~~	~~0.200E+02~~	~~13.95~~	~~RAJ~~
Total Activity :			2.234E+05	2.234E+05			11-16-98

Grand Total Activity : 2.234E+05 2.234E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2819-DUP

Acquisition date : 13-NOV-1998 15:54:17

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	107126.	477.59	6.0974E+02
NA-24	61.	1368.55	3.7355E+00
K-40	1717.	1460.75	1.7337E+02
AR-41	97.	1293.60	6.8058E+00
SC-46	325.	1120.55	6.7935E+00
CR-51	83286.	320.08	3.8277E+02
MN-54	241.	834.83	4.8493E+00
CO-56	242.	846.76	4.9066E+00
CO-58	229.	810.78	4.6570E+00
FE-59	208.	1099.25	9.5732E+00
CO-60	78.	1332.50	3.8728E+00
ZN-65	178.	1115.55	9.9779E+00
SE-75	80413.	264.66	5.4986E+01
KR-85	53127.	514.00	1.0935E+04
SR-85	53119.	514.01	4.9462E+01
Y-88	24.	1836.06	3.0348E+00
Y-91	163.	1204.67	1.7037E+03
NB-94	209.	871.09	4.6593E+00
NB-95	256.	765.78	4.7088E+00
ZR-95	278.	756.73	8.8883E+00
ZRNB-95	310.	724.18	2.2706E+01
RU-103	65400.	497.08	5.6211E+01
RURH-106	16797.	621.93	6.3378E+02
AG-108m	308.	722.94	5.4207E+00
AG-110M	110856.	657.76	8.9035E+01
SN-113	96584.	391.69	7.8011E+01
TE-123m	80047.	159.00	2.7860E+01
SB-124	17181.	602.73	3.1701E+01
SB-125	113509.	427.89	1.9970E+02
TE-125m	70023.	109.27	7.1570E+03
I-129	56676.	39.60	1.7085E+02
I-131	88968.	364.48	5.5470E+01
XE-131m	81312.	163.93	1.2221E+03
BA-133	87511.	356.02	7.0106E+01
CS-134	17091.	604.70	3.1749E+01
CS-136	250.	818.51	4.8918E+00
CS-138	54.	1435.86	1.6525E+01
CE-139	81764.	165.85	3.0124E+01
BA-140	37814.	537.31	1.7127E+02
LA-140	31.	1596.21	3.1337E+00
CE-141	76867.	145.44	4.6171E+01
CE-144	75695.	133.51	1.9450E+02
CEPR-144	75692.	133.51	3.8863E+02
EU-152	75675.	121.78	7.4671E+01
EU-154	111.	1274.51	1.2866E+01
EU-155	66060.	86.54	5.9741E+01
HF-181	84914.	482.18	7.0430E+01
TA-182	60737.	67.75	4.5657E+01
HG-203	80912.	279.20	4.1424E+01
BI-207	21778.	569.70	3.4010E+01
TL-208	80832.	277.36	5.3112E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2819-DUP

Acquisition date : 13-NOV-1998 15:54:17

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	56425.	46.50	3.8668E+02
BI-212	343.	727.18	7.8328E+01
PB-212	82674.	238.63	6.8975E+01
BI-214	17562.	609.31	7.0545E+01
PB-214	87190.	351.92	5.2131E+02
RA-224	81667.	240.99	7.6680E+02
RA-226	131697.	186.10	9.2233E+02
AC-228	440.	911.21	2.5861E+01
TH-229	64217.	88.47	7.8164E+01
U-232	58059.	57.78	8.9439E+03
PA-233	83041.	312.17	9.7607E+01
UTH-233	81002.	245.34	2.7634E+04
PA-234M	186.	1001.03	7.4132E+02
TH-234	59732.	63.29	4.8715E+02
U-235	129066.	185.71	5.5517E+01
U-237	58445.	59.54	5.2756E+01
NP-238	179.	984.45	1.7179E+01
NP-239	67573.	106.12	8.7922E+01
PU-239	75074.	129.30	3.3284E+05
AM-241	58446.	59.54	5.0401E+01

*****
 * 222-S Laboratory Counting Room 13-NOV-1998 21:30:10.93 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T002820-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Newright

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1061.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:35:47.84 sec
 Dead Time: 3.7%

Verified by:

RJ Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 13-NOV-1998 18:53:54.58
 Decayed to: 13-NOV-1998 18:53:54.58
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:13:20.37

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.21	716	1.08	24.54	23	5 58.1			
0	21.76	5296	1.10	43.65	40	8 12.5			
8	29.87	26827	1.74	59.86	52	28 4.3	4.86E+02	I-129	489.
8	31.96	171994	1.07	64.05	52	28 0.6			
8	36.29	56411	1.68	72.69	52	28 1.6			
2	72.77*	3028	1.12	145.65	142	13 19.6	3.05E+00		
2	74.94*	6031	0.98	149.98	142	13 8.9			
0	84.59*	3654	1.08	169.29	166	8 18.4		AM-243	563.
0	87.60*	1418	1.14	175.30	173	7 41.6		TH-228	1.633E+04
								NP-237	647.
								CD-109	2.225E+03
0	661.65*	1161605	1.45	1323.40	1315	16 0.2		CS-137	2.532E+05
0	726.88*	91	2.48	1453.88	1448	11 71.3			
0	1323.39	53	1.47	2647.16	2641	14 87.2			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002820-SAM

Acquisition date : 13-NOV-1998 18:53:54

Total number of lines in spectrum 12
 Number of unidentified lines 6
 Number of lines tentatively identified by NID 6 50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CD-109~~	~~462.90D~~	~~1.00~~	~~2.224E+03~~	~~2.225E+03~~	~~0.926E+03~~	~~41.62~~	~~NA~~
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~4.888E+02~~	~~4.888E+02~~	~~0.211E+02~~	~~4.31~~	~~NA~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~2.532E+05~~	~~2.532E+05~~	~~0.005E+05~~	~~0.19~~	~~NA~~
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~1.633E+04~~	~~1.633E+04~~	~~0.301E+04~~	~~18.40~~	~~NA~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~6.474E+02~~	~~6.474E+02~~	~~2.695E+02~~	~~41.62~~	~~NA~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~5.634E+02~~	~~5.634E+02~~	~~0.503E+02~~	~~8.93~~	~~NA~~
Total Activity :			2.734E+05	2.734E+05			11-16-98

Grand Total Activity : 2.734E+05 2.734E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002820-SAM

Acquisition date : 13-NOV-1998 18:53:54

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	36235.	477.59	1.1886E+03
NA-24	50.	1368.55	1.1634E+01
K-40	1775.	1460.75	5.9977E+02
AR-41	83.	1293.60	2.1156E+01
SC-46	262.	1120.55	2.1104E+01
CR-51	28042.	320.08	7.3602E+02
MN-54	168.	834.83	1.4004E+01
CO-56	171.	846.76	1.4300E+01
CO-58	170.	810.78	1.3878E+01
FE-59	139.	1099.25	2.7223E+01
CO-60	81.	1332.50	1.3457E+01
ZN-65	145.	1115.55	3.1267E+01
SE-75	27912.	264.66	1.0750E+02
KR-85	18457.	514.00	2.1691E+04
SR-85	18456.	514.01	9.8116E+01
Y-88	22.	1836.06	9.7394E+00
Y-91	129.	1204.67	5.2391E+03
NB-94	187.	871.09	1.5231E+01
NB-95	198.	765.78	1.4280E+01
ZR-95	183.	756.73	2.4936E+01
ZRNB-95	240.	724.18	6.8605E+01
RU-103	22679.	497.08	1.1119E+02
RURH-106	6049.	621.93	1.2945E+03
AG-108m	230.	722.94	1.6103E+01
AG-110M	32501.	657.76	1.6432E+02
SN-113	33069.	391.69	1.5182E+02
TE-123m	31710.	159.00	5.9614E+01
SB-124	6157.	602.73	6.4481E+01
SB-125	39566.	427.89	3.9329E+02
TE-125m	27136.	109.27	1.5465E+04
I-131	30602.	364.48	1.0798E+02
XE-131m	32518.	163.93	2.6216E+03
BA-133	30205.	356.02	1.3667E+02
CS-134	6200.	604.70	6.4984E+01
CS-136	163.	818.51	1.3707E+01
CS-138	53.	1435.86	5.3213E+01
CE-139	32698.	165.85	6.4583E+01
BA-140	13323.	537.31	3.4296E+02
LA-140	42.	1596.21	1.2132E+01
CE-141	30482.	145.44	9.9427E+01
CE-144	29633.	133.51	4.1843E+02
CEPR-144	29634.	133.51	8.3612E+02
EU-152	30040.	121.78	1.6259E+02
EU-154	97.	1274.51	4.1388E+01
EU-155	27468.	86.54	1.3347E+02
HF-181	29218.	482.18	1.3855E+02
TA-182	23414.	67.75	9.5323E+01
HG-203	27412.	279.20	7.9939E+01
BI-207	7706.	569.70	6.8514E+01
TL-208	27371.	277.36	1.0248E+03
PB-210	21004.	46.50	6.8357E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002820-SAM

Acquisition date : 13-NOV-1998 18:53:54

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	315.	727.18	2.5704E+02
PB-212	31827.	238.63	1.4238E+02
BI-214	6631.	609.31	1.4734E+02
PB-214	30035.	351.92	9.5864E+02
RA-224	30858.	240.99	1.5667E+03
RA-226	49096.	186.10	1.8953E+03
AC-228	413.	911.21	8.6431E+01
TH-229	26425.	88.47	1.7391E+02
U-232	21738.	57.78	1.7613E+04
PA-233	27898.	312.17	1.8746E+02
UTH-233	29858.	245.34	5.5772E+04
PA-234M	138.	1001.03	2.2233E+03
TH-234	22846.	63.29	9.9706E+02
U-235	48470.	185.71	1.1451E+02
U-237	22274.	59.54	1.0584E+02
NP-238	152.	984.45	5.4815E+01
NP-239	26191.	106.12	1.8995E+02
PU-239	30075.	129.30	7.2570E+05
AM-241	22274.	59.54	1.0115E+02

*****
 * 222-S Laboratory Counting Room 14-NOV-1998 03:42:29.08 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T002820-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

new light

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEAL
 File Number: dka300:[spec.GEAL]1g1065.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:35:44.03 sec
 Dead Time: 3.7%

Verified by:

RJ Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-NOV-1998 01:06:16.65
 Decayed to: 14-NOV-1998 01:06:16.65
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEAL]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:13:20.37

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.28	631	0.69	24.69	23	5 65.2			
0	21.77	5284	1.09	43.66	40	8 12.6			
10	30.24	27763	2.11	60.59	55	14 4.8	3.27E+02	I-129	506.
10	31.96	158803	1.00	64.05	55	14 0.7			
0	36.39	45368	1.12	72.91	69	10 2.1			
2	72.81*	2796	0.91	145.72	142	13 16.3	2.86E+00		
2	74.95*	5848	0.93	150.01	142	13 9.2		AM-243	546.
0	84.60*	3393	1.09	169.31	166	8 19.8		TH-228	1.516E+04
0	87.63*	1566	1.21	175.36	173	7 37.3		NP-237	715.
								CD-109	2.457E+03
0	303.07	565	1.14	606.22	604	6 83.8			
0	661.65*	1151082	1.45	1323.39	1315	16 0.2		CS-137	2.509E+05
0	1323.42	50	1.07	2647.23	2639	14 93.4			
0	1730.94	27	3.26	3462.61	3453	16101.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002820-DUP

Acquisition date : 14-NOV-1998 01:06:16

Total number of lines in spectrum 13
 Number of unidentified lines 7
 Number of lines tentatively identified by NID 6 46.15%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CD-109~~	~~462.90D~~	~~1.00~~	~~2.456E+03~~	~~2.457E+03~~	~~0.916E+03~~	~~37.30~~	~~raj~~
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~5.059E+02~~	~~5.059E+02~~	~~0.244E+02~~	~~4.83~~	~~raj~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~2.509E+05~~	~~2.509E+05~~	~~0.005E+05~~	~~0.19~~	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~1.516E+04~~	~~1.516E+04~~	~~0.300E+04~~	~~19.82~~	~~raj~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~7.149E+02~~	~~7.149E+02~~	~~2.666E+02~~	~~37.30~~	~~raj~~
~~AM-243~~	~~7380.00Y*****~~		~~5.463E+02~~	~~5.463E+02~~	~~0.501E+02~~	~~9.17~~	~~raj~~
Total Activity :			2.703E+05	2.703E+05			11-16-98

Grand Total Activity : 2.703E+05 2.703E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002820-DUP

Acquisition date : 14-NOV-1998 01:06:16

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uci/L)
BE-7	36054.	477.59	1.1856E+03
NA-24	54.	1368.55	1.2066E+01
K-40	1764.	1460.75	5.9782E+02
AR-41	83.	1293.60	2.1192E+01
SC-46	269.	1120.55	2.1372E+01
CR-51	28180.	320.08	7.3783E+02
MN-54	186.	834.83	1.4704E+01
CO-56	166.	846.76	1.4080E+01
CO-58	194.	810.78	1.4782E+01
FE-59	144.	1099.25	2.7754E+01
CO-60	81.	1332.50	1.3441E+01
ZN-65	149.	1115.55	3.1612E+01
SE-75	27719.	264.66	1.0713E+02
KR-85	18491.	514.00	2.1711E+04
SR-85	18487.	514.01	9.8198E+01
Y-88	15.	1836.06	8.2809E+00
Y-91	144.	1204.67	5.5222E+03
NB-94	163.	871.09	1.4240E+01
NB-95	178.	765.78	1.3547E+01
ZR-95	204.	756.73	2.6259E+01
ZRNB-95	213.	724.18	6.4754E+01
RU-103	22715.	497.08	1.1128E+02
RURH-106	5908.	621.93	1.2794E+03
AG-108m	193.	722.94	1.4804E+01
AG-110M	32155.	657.76	1.6345E+02
SN-113	32894.	391.69	1.5142E+02
TE-123m	31494.	159.00	5.9411E+01
SB-124	6186.	602.73	6.4632E+01
SB-125	38837.	427.89	3.8966E+02
TE-125m	27304.	109.27	1.5512E+04
I-131	30507.	364.48	1.0781E+02
XE-131m	32245.	163.93	2.6106E+03
BA-133	29863.	356.02	1.3590E+02
CS-134	6188.	604.70	6.4925E+01
CS-136	166.	818.51	1.3814E+01
CS-138	58.	1435.86	5.5374E+01
CE-139	32345.	165.85	6.4235E+01
BA-140	13072.	537.31	3.3973E+02
LA-140	37.	1596.21	1.1440E+01
CE-141	30421.	145.44	9.9328E+01
CE-144	29292.	133.51	4.1603E+02
CEPR-144	29293.	133.51	8.3131E+02
EU-152	29647.	121.78	1.6152E+02
EU-154	87.	1274.51	3.9432E+01
EU-155	27209.	86.54	1.3284E+02
HF-181	29018.	482.18	1.3807E+02
TA-182	23128.	67.75	9.4742E+01
HG-203	27405.	279.20	7.9929E+01
BI-207	7832.	569.70	6.9070E+01
TL-208	27197.	277.36	1.0216E+03
PB-210	20657.	46.50	6.7792E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002820-DUP

Acquisition date : 14-NOV-1998 01:06:16

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	296.	727.18	2.4959E+02
PB-212	31260.	238.63	1.4111E+02
BI-214	6521.	609.31	1.4612E+02
PB-214	29582.	351.92	9.5104E+02
RA-224	30460.	240.99	1.5566E+03
RA-226	48206.	186.10	1.8780E+03
AC-228	357.	911.21	8.0541E+01
TH-229	25726.	88.47	1.7161E+02
U-232	21977.	57.78	1.7709E+04
PA-233	28186.	312.17	1.8842E+02
UTH-233	29095.	245.34	5.5058E+04
PA-234M	131.	1001.03	2.1690E+03
TH-234	22603.	63.29	9.9179E+02
U-235	47427.	185.71	1.1328E+02
U-237	22018.	59.54	1.0523E+02
NP-238	132.	984.45	5.1186E+01
NP-239	26453.	106.12	1.9090E+02
PU-239	29968.	129.30	7.2441E+05
AM-241	22018.	59.54	1.0057E+02

*****
 * 222-S Laboratory Counting Room 14-NOV-1998 06:23:38.29 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T002821-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLU

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1066.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:36:38.95 sec
 Dead Time: 4.2%

Verified by:

RAJAN 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-NOV-1998 03:46:29.69
 Decayed to: 14-NOV-1998 03:46:29.69
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:13:20.37

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	9.26	1692	1.36	18.64	18	5 24.6			
0	12.09	1194	0.94	24.30	22	7 48.4			
0	21.77	6203	1.06	43.65	40	8 11.6			
10	30.24	38467	2.11	60.59	52	28 3.9	5.07E+02	I-129	701.
10	31.98	195943	1.07	64.07	52	28 0.6			
10	36.30	65747	1.67	72.72	52	28 1.5			
2	72.78*	3534	1.11	145.66	142	12 18.0	2.59E+00		
2	74.96*	6683	0.92	150.03	142	12 8.7			
0	84.63*	4198	1.10	169.37	166	8 17.2			
0	87.28*	1319	0.97	174.66	173	7 48.0			
0	186.13*	715	1.17	372.36	370	6100.7			
0	661.64*	1351077	1.45	1323.39	1315	16 0.2			
0	1322.22	121	1.81	2644.84	2637	16 37.8			

Handwritten notes:
 39%
 Pb x-rays
 AM-243 624.
 TH-228 1.876E+04
 NP-237 602.
 CD-109 2.068E+03
 U-235 79.7
 RA-226 1.311E+03
 CS-137 2.945E+05

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002821-SAM

Acquisition date : 14-NOV-1998 03:46:29

Total number of lines in spectrum 13
 Number of unidentified lines 7
 Number of lines tentatively identified by NID 6 46.15%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~GD-109~~	~~462.90D~~	~~1.00~~	~~2.068E+03~~	~~2.068E+03~~	~~0.992E+03~~	~~47.98~~	~~RMF~~
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~7.009E+02~~	~~7.009E+02~~	~~0.272E+02~~	~~3.88~~	~~RMF~~
CS-137	30.00Y	1.00	2.945E+05	2.945E+05	0.005E+05	0.18	
~~RA-226~~	~~1600.00Y~~	~~1.00~~	~~1.311E+03~~	~~1.311E+03~~	~~1.321E+03~~	~~100.70~~	~~RMF~~
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~1.876E+04~~	~~1.876E+04~~	~~0.323E+04~~	~~17.24~~	~~RMF~~
~~U-235~~	~~7.04E+08Y~~	~~1.00~~	~~7.974E+01~~	~~7.974E+01~~	~~8.030E+01~~	~~100.70~~	~~RMF~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~6.019E+02~~	~~6.019E+02~~	~~2.888E+02~~	~~47.98~~	~~RMF~~
~~AM-243~~	~~7380.00Y~~	~~1.00~~	~~6.242E+02~~	~~6.242E+02~~	~~0.540E+02~~	~~8.65~~	~~RMF~~
Total Activity :			3.186E+05	3.186E+05			11-16-8

Grand Total Activity : 3.186E+05 3.186E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002821-SAM

Acquisition date : 14-NOV-1998 03:46:29

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	42123.	477.59	1.2812E+03
NA-24	51.	1368.55	1.1727E+01
K-40	1778.	1460.75	6.0022E+02
AR-41	96.	1293.60	2.2710E+01
SC-46	280.	1120.55	2.1789E+01
CR-51	32997.	320.08	7.9819E+02
MN-54	189.	834.83	1.4819E+01
CO-56	176.	846.76	1.4498E+01
CO-58	173.	810.78	1.3999E+01
FE-59	139.	1099.25	2.7236E+01
CO-60	57.	1332.50	1.1469E+01
ZN-65	152.	1115.55	3.1998E+01
SE-75	32418.	264.66	1.1582E+02
KR-85	22005.	514.00	2.3676E+04
SR-85	22002.	514.01	1.0709E+02
Y-88	15.	1836.06	8.1361E+00
Y-91	146.	1204.67	5.5689E+03
NB-94	142.	871.09	1.3327E+01
NB-95	193.	765.78	1.4088E+01
ZR-95	235.	756.73	2.8082E+01
ZRNB-95	232.	724.18	6.7442E+01
RU-103	26321.	497.08	1.1975E+02
RURH-106	6935.	621.93	1.3854E+03
AG-108m	233.	722.94	1.6181E+01
AG-110M	38389.	657.76	1.7854E+02
SN-113	38538.	391.69	1.6386E+02
TE-123m	36890.	159.00	6.4284E+01
SB-124	7245.	602.73	6.9905E+01
SB-125	45684.	427.89	4.2252E+02
TE-125m	31331.	109.27	1.6613E+04
I-131	35465.	364.48	1.1621E+02
XE-131m	37215.	163.93	2.8040E+03
BA-133	34556.	356.02	1.4616E+02
CS-134	7199.	604.70	6.9989E+01
CS-136	205.	818.51	1.5291E+01
CS-138	55.	1435.86	5.3984E+01
CE-139	37648.	165.85	6.9285E+01
BA-140	15312.	537.31	3.6755E+02
LA-140	36.	1596.21	1.1307E+01
CE-141	35354.	145.44	1.0705E+02
CE-144	34526.	133.51	4.5155E+02
CEPR-144	34526.	133.51	9.0228E+02
EU-152	34781.	121.78	1.7490E+02
EU-154	94.	1274.51	4.0720E+01
EU-155	31795.	86.54	1.4357E+02
HF-181	34117.	482.18	1.4967E+02
TA-182	27106.	67.75	1.0254E+02
HG-203	31833.	279.20	8.6123E+01
BI-207	8969.	569.70	7.3882E+01
TL-208	31789.	277.36	1.1042E+03
PB-210	24341.	46.50	7.3566E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002821-SAM

Acquisition date : 14-NOV-1998 03:46:29

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	254.	727.18	2.3153E+02
PB-212	36634.	238.63	1.5272E+02
BI-214	7643.	609.31	1.5811E+02
PB-214	35012.	351.92	1.0400E+03
RA-224	35729.	240.99	1.6855E+03
AC-228	394.	911.21	8.4544E+01
TH-229	30196.	88.47	1.8587E+02
U-232	25310.	57.78	1.8999E+04
PA-233	32489.	312.17	2.0225E+02
UTH-233	34755.	245.34	6.0158E+04
PA-234M	142.	1001.03	2.2554E+03
TH-234	26358.	63.29	1.0707E+03
U-237	25705.	59.54	1.1367E+02
NP-238	140.	984.45	5.2572E+01
NP-239	30272.	106.12	2.0418E+02
PU-239	34873.	129.30	7.8127E+05
AM-241	25705.	59.54	1.0863E+02

*****
 * 222-S Laboratory Counting Room 14-NOV-1998 09:10:39.10 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2821-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SL Usher

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1067.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:36:43.08 sec
 Dead Time: 4.3%

Verified by:

Raymer 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-NOV-1998 06:33:27.32
 Decayed to: 14-NOV-1998 06:33:27.32
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:13:20.37

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	9.23	8892	1.39	18.59	16	7	5.0			
0	12.00	830	0.73	24.12	22	7	69.5			
0	21.84	6141	1.11	43.81	40	8	11.7			
10	30.24	39878	2.11	60.59	52	27	3.8	5.64E+02	I-129	727.
10	31.98	198462	1.06	64.07	52	27	0.6			
10	36.30	66172	1.68	72.71	52	27	1.5			
2	72.75*	3674	1.10	145.61	142	13	17.4	4.55E-01		
2	74.95*	6897	0.96	150.01	142	13	8.4		AM-243	644.
0	84.73*	4678	1.14	169.57	166	7	13.9		TH-228	2.090E+04
0	87.39*	1181	0.88	174.89	173	7	54.0		NP-237	539.
									CD-109	1.852E+03
0	186.69*	1231	1.58	373.46	371	6	59.2		U-235	137.
									RA-226	2.258E+03
0	661.64*	1364326	1.45	1323.38	1315	16	0.2		CS-137	2.974E+05
0	1322.58	115	1.75	2645.55	2638	16	40.1			
0	1764.47*	35	2.27	3529.69	3522	17	101.5			
0	1848.59	21	1.30	3698.03	3692	12	93.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2821-DUP

Acquisition date : 14-NOV-1998 06:33:27

Total number of lines in spectrum 15
 Number of unidentified lines 8
 Number of lines tentatively identified by NID 7 46.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CD-109~~	~~462.90D~~	~~1.00~~	~~1.852E+03~~	~~1.852E+03~~	~~0.999E+03~~	~~53.96~~	~~K~~
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~7.266E+02~~	~~7.266E+02~~	~~0.274E+02~~	~~3.77~~	~~M~~
CS-137	30.00Y	1.00	2.974E+05	2.974E+05	0.005E+05	0.18	
RA-226	1600.00Y*****	1.00	2.258E+03	2.258E+03	1.336E+03	59.16	K
TH-228	1.91Y	1.00	2.090E+04	2.090E+04	0.291E+04	13.90	K
U-235	7.04E+08Y	1.00	1.373E+02	1.373E+02	0.812E+02	59.16	K
NP-237	2.14E+06Y	1.00	5.389E+02	5.389E+02	2.908E+02	53.96	K
AM-243	7380.00Y*****	1.00	6.443E+02	6.443E+02	0.544E+02	8.44	K
Total Activity :			3.244E+05	3.244E+05			11-16-98

Grand Total Activity : 3.244E+05 3.244E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2821-DUP

Acquisition date : 14-NOV-1998 06:33:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	42850.	477.59	1.2922E+03
NA-24	56.	1368.55	1.2260E+01
K-40	1746.	1460.75	5.9486E+02
AR-41	90.	1293.60	2.2071E+01
SC-46	252.	1120.55	2.0707E+01
CR-51	33395.	320.08	8.0298E+02
MN-54	211.	834.83	1.5640E+01
CO-56	161.	846.76	1.3902E+01
CO-58	191.	810.78	1.4680E+01
FE-59	131.	1099.25	2.6512E+01
CO-60	78.	1332.50	1.3227E+01
ZN-65	155.	1115.55	3.2232E+01
SE-75	32739.	264.66	1.1639E+02
KR-85	21961.	514.00	2.3652E+04
SR-85	21957.	514.01	1.0698E+02
Y-88	24.	1836.06	1.0035E+01
Y-91	140.	1204.67	5.4491E+03
NB-94	168.	871.09	1.4469E+01
NB-95	197.	765.78	1.4225E+01
ZR-95	218.	756.73	2.7096E+01
ZRNB-95	249.	724.18	6.9810E+01
RU-103	26648.	497.08	1.2049E+02
RURH-106	7067.	621.93	1.3984E+03
AG-108m	238.	722.94	1.6352E+01
AG-110M	39269.	657.76	1.8057E+02
SN-113	38867.	391.69	1.6455E+02
TE-123m	37460.	159.00	6.4778E+01
SB-124	7071.	602.73	6.9069E+01
SB-125	46167.	427.89	4.2474E+02
TE-125m	32021.	109.27	1.6795E+04
I-131	35810.	364.48	1.1678E+02
XE-131m	37670.	163.93	2.8210E+03
BA-133	35345.	356.02	1.4781E+02
CS-134	7214.	604.70	7.0059E+01
CS-136	177.	818.51	1.4238E+01
CS-138	64.	1435.86	5.7967E+01
CE-139	37899.	165.85	6.9514E+01
BA-140	15576.	537.31	3.7069E+02
LA-140	45.	1596.21	1.2581E+01
CE-141	35523.	145.44	1.0731E+02
CE-144	34725.	133.51	4.5284E+02
CEPR-144	34723.	133.51	9.0484E+02
EU-152	35016.	121.78	1.7549E+02
EU-154	96.	1274.51	4.1254E+01
EU-155	32073.	86.54	1.4419E+02
HF-181	34064.	482.18	1.4956E+02
TA-182	27536.	67.75	1.0334E+02
HG-203	32158.	279.20	8.6560E+01
BI-207	9147.	569.70	7.4606E+01
TL-208	31978.	277.36	1.1074E+03
PB-210	24587.	46.50	7.3936E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2821-DUP

Acquisition date : 14-NOV-1998 06:33:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	313.	727.18	2.5617E+02
PB-212	36840.	238.63	1.5314E+02
BI-214	7593.	609.31	1.5759E+02
PB-214	35444.	351.92	1.0468E+03
RA-224	36290.	240.99	1.6986E+03
AC-228	379.	911.21	8.2888E+01
TH-229	30310.	88.47	1.8622E+02
U-232	25691.	57.78	1.9141E+04
PA-233	32928.	312.17	2.0361E+02
UTH-233	34883.	245.34	6.0268E+04
PA-234M	137.	1001.03	2.2133E+03
TH-234	26673.	63.29	1.0770E+03
U-237	26049.	59.54	1.1443E+02
NP-238	139.	984.45	5.2463E+01
NP-239	30144.	106.12	2.0376E+02
PU-239	34605.	129.30	7.7826E+05
AM-241	26049.	59.54	1.0935E+02

*****
 * 222-S Laboratory Counting Room 14-NOV-1998 16:28:36.86 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2939-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1069.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:46:38.73 sec
 Dead Time: 10%

Verified by:

PT Jones 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-NOV-1998 13:41:30.05
 Decayed to: 14-NOV-1998 13:41:30.05
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	9.34	4721	1.42	18.81	18	5 14.0			
1	11.48	408	0.69	23.09	22	7 88.1	4.46E+01		
1	12.51	2784	0.82	25.15	22	7 25.2			
0	21.91	16985	1.19	43.94	41	7 6.1			
0	31.97	561598	1.15	64.05	60	9 0.4			
0	36.47	155046	1.13	73.05	69	9 1.0			
2	72.84*	3127	1.12	145.79	143	11 23.7	1.73E+00		
2	75.03*	6407	1.09	150.18	143	11 15.2		AM-243	175.
0	84.66*	2906	1.17	169.42	168	6 29.6		TH-228	3.759E+03
0	190.30	71584	8.66	380.69	369	28 5.8			
0	661.65*	3638885	1.54	1323.40	1315	17 0.1		CS-137	2.330E+05
0	692.47	1158	2.72	1385.05	1377	17 22.9			
0	1322.89	641	2.19	2646.17	2637	20 12.1			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2939-SAM

Acquisition date : 14-NOV-1998 13:41:30

Total number of lines in spectrum 13
 Number of unidentified lines 10
 Number of lines tentatively identified by NID 3 23.08%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.330E+05	2.330E+05	0.003E+05	0.11	
TH-228	1.91Y	1.00	3.758E+03	3.759E+03	1.113E+03	29.61	
AM-243	7380.00Y*****		1.753E+02	1.753E+02	0.266E+02	15.17	
Total Activity :			2.369E+05	2.369E+05			

11-16-98

Grand Total Activity : 2.369E+05 2.369E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2939-SAM

Acquisition date : 14-NOV-1998 13:41:30

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	113334.	477.59	6.2713E+02
NA-24	60.	1368.55	3.7212E+00
K-40	1729.	1460.75	1.7393E+02
AR-41	85.	1293.60	6.4012E+00
SC-46	334.	1120.55	6.8817E+00
CR-51	88364.	320.08	3.9425E+02
MN-54	248.	834.83	4.9159E+00
CO-56	248.	846.76	4.9640E+00
CO-58	210.	810.78	4.4702E+00
FE-59	193.	1099.25	9.2283E+00
CO-60	86.	1332.50	4.0394E+00
ZN-65	206.	1115.55	1.0725E+01
SE-75	85306.	264.66	5.6631E+01
KR-85	56810.	514.00	1.1307E+04
SR-85	56806.	514.01	5.1145E+01
Y-88	22.	1836.06	2.9041E+00
Y-91	153.	1204.67	1.6510E+03
NB-94	242.	871.09	5.0047E+00
NB-95	268.	765.78	4.8112E+00
ZR-95	238.	756.73	8.2426E+00
ZRNB-95	307.	724.18	2.2588E+01
RU-103	69061.	497.08	5.7759E+01
RURH-106	17603.	621.93	6.4874E+02
AG-108m	315.	722.94	5.4819E+00
CD-109	68883.	88.03	5.5398E+02
AG-110M	121835.	657.76	9.3333E+01
SN-113	101874.	391.69	8.0115E+01
TE-123m	84977.	159.00	2.8703E+01
SB-124	18446.	602.73	3.2843E+01
SB-125	120247.	427.89	2.0553E+02
TE-125m	74214.	109.27	7.3676E+03
I-129	60079.	39.60	1.7589E+02
I-131	94066.	364.48	5.7036E+01
XE-131m	86645.	163.93	1.2614E+03
BA-133	92896.	356.02	7.2226E+01
CS-134	18378.	604.70	3.2918E+01
CS-136	241.	818.51	4.8134E+00
CS-138	59.	1435.86	1.7356E+01
CE-139	87297.	165.85	3.1124E+01
BA-140	39947.	537.31	1.7602E+02
LA-140	32.	1596.21	3.1615E+00
CE-141	81408.	145.44	4.7513E+01
CE-144	80028.	133.51	1.9997E+02
CEPR-144	80027.	133.51	3.9958E+02
EU-152	80029.	121.78	7.6784E+01
EU-154	99.	1274.51	1.2191E+01
EU-155	70036.	86.54	6.1509E+01
HF-181	89999.	482.18	7.2504E+01
TA-182	64239.	67.75	4.6952E+01
HG-203	85560.	279.20	4.2595E+01
BI-207	23336.	569.70	3.5201E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2939-SAM

Acquisition date : 14-NOV-1998 13:41:30

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	86317.	277.36	5.4881E+02
PB-210	60017.	46.50	3.9878E+02
BI-212	354.	727.18	7.9480E+01
PB-212	87081.	238.63	7.0786E+01
BI-214	18354.	609.31	7.2111E+01
PB-214	92318.	351.92	5.3944E+02
RA-224	86615.	240.99	7.8970E+02
RA-226	140194.	186.10	9.5157E+02
AC-228	427.	911.21	2.5475E+01
TH-229	68623.	88.47	8.0795E+01
U-232	61484.	57.78	9.2033E+03
PA-233	87948.	312.17	1.0045E+02
UTH-233	86301.	245.34	2.8522E+04
PA-234M	196.	1001.03	7.6070E+02
TH-234	62914.	63.29	4.9993E+02
U-235	137376.	185.71	5.7274E+01
NP-237	70230.	86.48	1.6293E+02
U-237	62449.	59.54	5.4531E+01
NP-238	197.	984.45	1.7972E+01
NP-239	71331.	106.12	9.0338E+01
PU-239	80178.	129.30	3.4395E+05
AM-241	62450.	59.54	5.2095E+01

*****
 * 222-S Laboratory Counting Room 14-NOV-1998 19:37:14.18 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26737
 Sample ID: S98T2939-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

[Signature]

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1070.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:46:11.22 sec
 Dead Time: 9.7%

Verified by:

[Signature] 11-16-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-NOV-1998 16:50:33.61
 Decayed to: 14-NOV-1998 16:50:33.61
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:12:15.61

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	9.41	4782	1.40	18.95	18	5 13.6			
0	12.60	1430	0.53	25.33	23	6 55.8			
0	21.83	18581	1.29	43.78	40	8 6.1			
0	31.95	546190	1.14	64.02	60	9 0.4			
0	36.45	151238	1.13	73.01	69	9 1.0			
2	72.92*	2850	1.12	145.94	143	11 33.5	1.13E+00		
2	74.98*	5876	1.06	150.08	143	11 14.5	Pb-215	AM-243	161.
0	84.65*	2831	1.07	169.42	168	6 30.1		TH-228	3.662E+03
0	190.19	51374	7.98	380.47	370	25 7.3			
0	661.66*	3545044	1.54	1323.42	1315	17 0.1		CS-137	2.270E+05
0	692.44	1158	2.75	1384.99	1378	18 22.8			
0	1323.04	638	1.94	2646.48	2638	19 11.1			
0	1590.21	90	5.97	3181.02	3168	24 41.0			

Summary of Nuclide Activity

Sample ID : S98T2939-DUP

Page : 2

Acquisition date : 14-NOV-1998 16:50:33

Total number of lines in spectrum 13
 Number of unidentified lines 10
 Number of lines tentatively identified by NID 3 23.08%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.270E+05	2.270E+05	0.002E+05	0.11	
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~3.662E+03~~	~~3.662E+03~~	~~1.102E+03~~	~~30.10~~	~~RAJ~~
~~AM-243~~	~~7380.00Y~~	~~1.00~~	~~1.608E+02~~	~~1.608E+02~~	~~0.233E+02~~	~~14.50~~	~~RAJ~~
Total Activity :			2.308E+05	2.308E+05			11-16-98

Grand Total Activity : 2.308E+05 2.308E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : S98T2939-DUP

Page : 3

Acquisition date : 14-NOV-1998 16:50:33

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	110487.	477.59	6.1922E+02
NA-24	55.	1368.55	3.5796E+00
K-40	1745.	1460.75	1.7476E+02
AR-41	92.	1293.60	6.6358E+00
SC-46	317.	1120.55	6.7071E+00
CR-51	86070.	320.08	3.8910E+02
MN-54	257.	834.83	5.0025E+00
CO-56	227.	846.76	4.7599E+00
CO-58	227.	810.78	4.6380E+00
FE-59	187.	1099.25	9.0897E+00
CO-60	94.	1332.50	4.2263E+00
ZN-65	195.	1115.55	1.0433E+01
SE-75	83394.	264.66	5.5994E+01
KR-85	55166.	514.00	1.1142E+04
SR-85	55158.	514.01	5.0400E+01
Y-88	22.	1836.06	2.9041E+00
Y-91	147.	1204.67	1.6202E+03
NB-94	245.	871.09	5.0283E+00
NB-95	256.	765.78	4.7095E+00
ZR-95	267.	756.73	8.7180E+00
ZRNB-95	292.	724.18	2.2051E+01
RU-103	67997.	497.08	5.7313E+01
RURH-106	17096.	621.93	6.3938E+02
AG-108m	295.	722.94	5.3112E+00
CD-109	66775.	88.03	5.4546E+02
AG-110M	116104.	657.76	9.1115E+01
SN-113	99404.	391.69	7.9140E+01
TE-123m	82682.	159.00	2.8314E+01
SB-124	18032.	602.73	3.2473E+01
SB-125	117278.	427.89	2.0298E+02
TE-125m	72216.	109.27	7.2679E+03
I-129	57967.	39.60	1.7278E+02
I-131	91689.	364.48	5.6311E+01
XE-131m	84320.	163.93	1.2444E+03
BA-133	90322.	356.02	7.1220E+01
CS-134	17973.	604.70	3.2554E+01
CS-136	241.	818.51	4.8085E+00
CS-138	55.	1435.86	1.6716E+01
CE-139	85123.	165.85	3.0735E+01
BA-140	39051.	537.31	1.7404E+02
LA-140	47.	1596.21	3.7972E+00
CE-141	79943.	145.44	4.7084E+01
CE-144	78270.	133.51	1.9777E+02
CEPR-144	78268.	133.51	3.9518E+02
EU-152	77826.	121.78	7.5722E+01
EU-154	100.	1274.51	1.2208E+01
EU-155	68074.	86.54	6.0643E+01
HF-181	88026.	482.18	7.1706E+01
TA-182	62741.	67.75	4.6402E+01
HG-203	83717.	279.20	4.2134E+01
BI-207	22476.	569.70	3.4549E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2939-DUP

Acquisition date : 14-NOV-1998 16:50:33

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	83497.	277.36	5.3979E+02
PB-210	58428.	46.50	3.9348E+02
BI-212	346.	727.18	7.8591E+01
PB-212	85394.	238.63	7.0098E+01
BI-214	17917.	609.31	7.1252E+01
PB-214	90047.	351.92	5.3139E+02
RA-224	84333.	240.99	7.7922E+02
RA-226	136893.	186.10	9.4032E+02
AC-228	438.	911.21	2.5813E+01
TH-229	66524.	88.47	7.9553E+01
U-232	60144.	57.78	9.1027E+03
PA-233	85458.	312.17	9.9015E+01
UTH-233	83989.	245.34	2.8138E+04
PA-234M	183.	1001.03	7.3566E+02
TH-234	61882.	63.29	4.9582E+02
U-235	133953.	185.71	5.6557E+01
NP-237	68278.	86.48	1.6065E+02
U-237	60781.	59.54	5.3799E+01
NP-238	174.	984.45	1.6932E+01
NP-239	69326.	106.12	8.9057E+01
PU-239	77549.	129.30	3.3827E+05
AM-241	60780.	59.54	5.1396E+01

LABCORE Completed Worklist Report for Worklist# 26738

Analyst: gl Instrument: GEA02

Book#: 28B57

Method: LA-548-121 Rev/Mod F-P

Worklist Comment: BY-105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	GEA-01	CO60-02	LIQUID	4.522e-03	4.99e-03	110.349 % Recovery
1 STD	0	0	GEA-01	CO60-02R	LIQUID	1	3.08	3.080 % Ct Error
1 STD	0	0	GEA-01	CS13702	LIQUID	7.451e-03	7.61e-03	102.134 % Recovery
1 STD	0	0	GEA-01	CS13702R	LIQUID	1	3.07	3.070 % Ct Error
2 BLNK	0	0	GEA-01	CO60-02	LIQUID	1	<8.53e-4	uCi/mL
2 BLNK	0	0	GEA-01	CS13702	LIQUID	1	1.04e-03	0.001 uCi/mL
3 SAMPLE	S98T002960	0	GEA-01	CO60-02	LIQUID	N/A	< 3.339e-03	3.34e-003 uCi/mL
3 SAMPLE	S98T002960	0	GEA-01	CO60-02R	LIQUID	N/A	n/a	% Ct. Error
3 SAMPLE	S98T002960	0	GEA-01	CS13702	LIQUID	N/A	1.730e+02	uCi/mL
3 SAMPLE	S98T002960	0	GEA-01	CS13702R	LIQUID	N/A	0.140	% Ct. Error
4 DUP	S98T002960	0	GEA-01	CO60-02	LIQUID	<3.34e-3	<3.25e-3	RFD
4 DUP	S98T002960	0	GEA-01	CO60-02R	LIQUID	1	n/a	% Ct Error
4 DUP	S98T002960	0	GEA-01	CS13702	LIQUID	1.73e+02	1.71e+02	1.163 RFD
4 DUP	S98T002960	0	GEA-01	CS13702R	LIQUID	1	0.140	% Ct Error
5 SAMPLE	S98T002961	0	GEA-01	CO60-02	LIQUID	N/A	< 3.239e-03	3.24e-003 uCi/mL
5 SAMPLE	S98T002961	0	GEA-01	CO60-02R	LIQUID	N/A	n/a	% Ct. Error
5 SAMPLE	S98T002961	0	GEA-01	CS13702	LIQUID	N/A	1.830e+02	uCi/mL
5 SAMPLE	S98T002961	0	GEA-01	CS13702R	LIQUID	N/A	0.130	% Ct. Error
6 DUP	S98T002961	0	GEA-01	CO60-02	LIQUID	<3.24e-3	<3.32e-3	RFD
6 DUP	S98T002961	0	GEA-01	CO60-02R	LIQUID	1	n/a	% Ct Error
6 DUP	S98T002961	0	GEA-01	CS13702	LIQUID	1.83e+02	1.84e+02	0.545 RFD
6 DUP	S98T002961	0	GEA-01	CS13702R	LIQUID	1	0.130	% Ct Error
7 SAMPLE	S98T002962	0	GEA-01	CO60-02	LIQUID	N/A	< 3.438e-03	3.44e-003 uCi/mL
7 SAMPLE	S98T002962	0	GEA-01	CO60-02R	LIQUID	N/A	n/a	% Ct. Error
7 SAMPLE	S98T002962	0	GEA-01	CS13702	LIQUID	N/A	2.050e+02	uCi/mL
7 SAMPLE	S98T002962	0	GEA-01	CS13702R	LIQUID	N/A	0.130	% Ct. Error
8 DUP	S98T002962	0	GEA-01	CO60-02	LIQUID	<3.44e-3	<3.14e-3	RFD
8 DUP	S98T002962	0	GEA-01	CO60-02R	LIQUID	1	n/a	% Ct Error
8 DUP	S98T002962	0	GEA-01	CS13702	LIQUID	2.05e+02	2.05e+02	0.000 RFD
8 DUP	S98T002962	0	GEA-01	CS13702R	LIQUID	1	0.130	% Ct Error
9 SAMPLE	S98T002963	0	GEA-01	CO60-02	LIQUID	N/A	< 3.391e-03	3.39e-003 uCi/mL
9 SAMPLE	S98T002963	0	GEA-01	CO60-02R	LIQUID	N/A	n/a	% Ct. Error
9 SAMPLE	S98T002963	0	GEA-01	CS13702	LIQUID	N/A	2.360e+02	uCi/mL
9 SAMPLE	S98T002963	0	GEA-01	CS13702R	LIQUID	N/A	0.120	% Ct. Error
10 DUP	S98T002963	0	GEA-01	CO60-02	LIQUID	<3.39e-3	<3.44e-3	RFD
10 DUP	S98T002963	0	GEA-01	CO60-02R	LIQUID	1	n/a	% Ct Error
10 DUP	S98T002963	0	GEA-01	CS13702	LIQUID	2.36e+02	2.38e+02	0.844 RFD
10 DUP	S98T002963	0	GEA-01	CS13702R	LIQUID	1	0.120	% Ct Error
11 SAMPLE	S98T002964	0	GEA-01	CO60-02	LIQUID	N/A	< 2.975e-03	2.98e-003 uCi/mL
11 SAMPLE	S98T002964	0	GEA-01	CO60-02R	LIQUID	N/A	n/a	% Ct. Error
11 SAMPLE	S98T002964	0	GEA-01	CS13702	LIQUID	N/A	2.220e+02	uCi/mL

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26738

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
11	SAMPLE	S98T002964	0		@GEA-01	CS13702R LIQUID	N/A	0.120		% Ct. Error
12	DUP	S98T002964	0		@GEA-01	CO60-02 LIQUID	<2.98e-3	<3.35e-3		RPD
12	DUP	S98T002964	0		@GEA-01	CO60-02R LIQUID	1	n/a		% Ct Error
12	DUP	S98T002964	0		@GEA-01	CS13702 LIQUID	2.22e+02	2.22e+02	0.000	RPD
12	DUP	S98T002964	0		@GEA-01	CS13702R LIQUID	1	0.120	0.120	% Ct Error

Final page for worklist# 26738

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

11-19-98
_____
Reviewer Signature Date

*****

* 222-S Laboratory Counting Room 17-NOV-1998 14:58:50.79 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26738
 Sample ID: WL26738-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Sylvia Cichon

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2839.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:20.05 sec
 Dead Time: 0.7%

Verified by:

John Relyea 18 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 17-NOV-1998 14:07:57.23
 Decayed to: 17-NOV-1998 14:07:57.23
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.54*	5230	1.49	1323.08	1316	15	3.1		CS-137	7.61
0	1173.00	2490	1.86	2346.04	2337	19	4.5		CO-60	4.94
0	1332.21	2285	1.85	2664.51	2655	18	4.3		CO-60	5.03
0	1460.69*	8	1.97	2921.49	2915	13257.3			K-40	0.187

Summary of Nuclide Activity

Page : 2

Sample ID : WL26738-STD

Acquisition date : 17-NOV-1998 14:07:57

Total number of lines in spectrum	4	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	4	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.869E-01~~	~~1.869E-01~~	~~4.809E-01~~	~~257.26~~	~~2R~~
CO-60	5.27Y	1.00	4.987E+00	4.987E+00	0.154E+00	3.08	
CS-137	30.00Y	1.00	7.607E+00	7.607E+00	0.234E+00	3.07	
Total Activity :			1.278E+01	1.278E+01			

Grand Total Activity :	1.278E+01	1.278E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26738-STD

Acquisition date : 17-NOV-1998 14:07:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	260.	477.59	6.9414E-01
NA-24	12.	1368.55	4.3924E-02
AR-41	23.	1293.60	6.2583E-02
SC-46	106.	1120.55	9.6414E-02
CR-51	266.	320.08	4.9655E-01
MN-54	154.	834.83	9.0886E-02
CO-56	114.	846.76	7.9743E-02
CO-58	113.	810.78	7.7062E-02
FE-59	138.	1099.25	1.9095E-01
ZN-65	117.	1115.55	1.9877E-01
SE-75	340.	264.66	8.1710E-02
KR-85	221.	514.00	1.6394E+01
SR-85	221.	514.01	7.4116E-02
Y-88	2.	1836.06	2.7939E-02
Y-91	39.	1204.67	2.1372E+01
NB-94	143.	871.09	9.1016E-02
NB-95	120.	765.78	7.5300E-02
ZR-95	127.	756.73	1.4035E-01
ZRNB-95	109.	724.18	3.1516E-01
RU-103	215.	497.08	7.4692E-02
RURH-106	125.	621.93	1.2957E+00
AG-108m	106.	722.94	7.4423E-02
CD-109	272.	88.03	1.3296E+00
AG-110M	212.	657.76	9.1712E-02
SN-113	251.	391.69	9.1692E-02
TE-123m	340.	159.00	4.3248E-02
SB-124	128.	602.73	6.4722E-02
SB-125	288.	427.89	2.3168E-01
TE-125m	281.	109.27	1.3184E+01
I-129	292.	39.60	1.5843E+01
I-131	283.	364.48	7.1646E-02
XE-131m	345.	163.93	1.8776E+00
BA-133	269.	356.02	8.9306E-02
CS-134	117.	604.70	6.2210E-02
CS-136	112.	818.51	7.7295E-02
CS-138	7.	1435.86	7.6866E-02
CE-139	332.	165.85	4.5293E-02
BA-140	140.	537.31	2.4438E-01
LA-140	4.	1596.21	3.2835E-02
CE-141	333.	145.44	7.4210E-02
CE-144	316.	133.51	3.1846E-01
CEPR-144	316.	133.51	6.3633E-01
EU-152	317.	121.78	1.2883E-01
EU-154	28.	1274.51	1.6947E-01
EU-155	271.	86.54	1.5064E-01
HF-181	233.	482.18	8.5493E-02
TA-182	281.	67.75	2.1714E-01
HG-203	338.	279.20	6.1107E-02
BI-207	135.	569.70	6.2963E-02
TL-208	344.	277.36	7.9129E-01
PB-210	289.	46.50	1.0868E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26738-STD

Acquisition date : 17-NOV-1998 14:07:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	117.	727.18	1.0721E+00
PB-212	366.	238.63	1.0484E-01
BI-214	158.	609.31	1.5753E-01
PB-214	301.	351.92	2.8940E-01
RA-224	351.	240.99	1.1393E+00
RA-226	432.	186.10	1.2208E+00
AC-228	179.	911.21	3.9487E-01
TH-228	262.	84.37	4.0208E+00
TH-229	272.	88.47	1.9292E-01
U-232	267.	57.78	7.5847E+01
PA-233	270.	312.17	1.2770E-01
UTH-233	365.	245.34	4.2344E+01
PA-234M	126.	1001.03	1.4707E+01
TH-234	250.	63.29	2.7703E+00
U-235	430.	185.71	7.4042E-02
NP-237	271.	86.48	3.9889E-01
U-237	319.	59.54	4.2799E-01
NP-238	154.	984.45	3.7608E-01
NP-239	332.	106.12	1.8177E-01
PU-239	313.	129.30	5.5335E+02
AM-241	319.	59.54	4.1054E-01
AM-243	280.	74.67	1.1410E-01

*****
 * 222-S Laboratory Counting Room 17-NOV-1998 17:44:10.56 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: WL26738-BLNK
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

John Relyea

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2840.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:29.81 sec
 Dead Time: 0.3%

Verified by:

John Relyea 18 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 17-NOV-1998 15:13:11.80
 Decayed to: 17-NOV-1998 15:13:11.80
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	661.69*	49	1.25	1323.37	1318	10 93.9		CS-137	1.04
0	1377.08	20	1.53	2754.25	2748	11 77.6			
0	1460.64*	34	1.85	2921.41	2912	17130.0		K-40	11.2

Summary of Nuclide Activity

Page : 2

Sample ID : WL26738-BLNC

Acquisition date : 17-NOV-1998 15:13:11

Total number of lines in spectrum 3
 Number of unidentified lines 1
 Number of lines tentatively identified by NID 2 66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.118E+01~~	~~1.118E+01~~	~~1.453E+01~~	~~129.97~~	~~100~~
CS-137	30.00Y	1.00	1.045E+00	1.045E+00	0.981E+00	93.90	
Total Activity :			1.222E+01	1.222E+01			

Grand Total Activity : 1.222E+01 1.222E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : WL26738-BLNK

Page : 3

Acquisition date : 17-NOV-1998 15:13:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	103.	477.59	6.5539E+00
NA-24	24.	1368.55	8.9242E-01
AR-41	32.	1293.60	1.4296E+00
SC-46	57.	1120.55	1.0632E+00
CR-51	135.	320.08	5.2559E+00
MN-54	39.	834.83	6.9776E-01
CO-56	40.	846.76	7.1731E-01
CO-58	37.	810.78	6.7217E-01
FE-59	35.	1099.25	1.4692E+00
CO-60	26.	1332.50	8.5322E-01
ZN-65	36.	1115.55	1.6854E+00
SE-75	168.	264.66	8.5246E-01
KR-85	105.	514.00	1.6864E+02
SR-85	105.	514.01	7.6201E-01
Y-88	9.	1836.06	7.3558E-01
Y-91	37.	1204.67	3.0842E+02
NB-94	50.	871.09	8.1471E-01
NB-95	62.	765.78	8.1135E-01
ZR-95	54.	756.73	1.3776E+00
ZRNB-95	64.	724.18	3.5898E+00
RU-103	78.	497.08	6.7693E-01
RURH-106	62.	621.93	1.3687E+01
AG-108m	67.	722.94	8.8221E-01
CD-109	202.	88.03	1.6920E+01
AG-110M	58.	657.76	7.3159E-01
SN-113	111.	391.69	9.0950E-01
TE-123m	218.	159.00	5.1182E-01
SB-124	73.	602.73	7.2619E-01
SB-125	103.	427.89	2.0773E+00
TE-125m	188.	109.27	1.5921E+02
I-129	136.	39.60	1.6130E+02
I-131	125.	364.48	7.1303E-01
XE-131m	198.	163.93	2.1119E+01
BA-133	126.	356.02	9.0931E-01
CS-134	78.	604.70	7.5437E-01
CS-136	45.	818.51	7.3810E-01
CS-138	11.	1435.86	2.7455E+00
CE-139	180.	165.85	4.9512E-01
BA-140	67.	537.31	2.5299E+00
LA-140	7.	1596.21	5.9513E-01
CE-141	189.	145.44	8.2939E-01
CE-144	217.	133.51	3.8905E+00
CEPR-144	217.	133.51	7.7733E+00
EU-152	186.	121.78	1.4636E+00
EU-154	27.	1274.51	2.4131E+00
EU-155	204.	86.54	1.9270E+00
HF-181	81.	482.18	7.5895E-01
TA-182	168.	67.75	2.4925E+00
HG-203	173.	279.20	6.4995E-01
BI-207	87.	569.70	7.4978E-01
TL-208	191.	277.36	8.7443E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26738-BLNK

Acquisition date : 17-NOV-1998 15:13:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	157.	46.50	1.1897E+02
BI-212	74.	727.18	1.2637E+01
PB-212	212.	238.63	1.1818E+00
BI-214	122.	609.31	2.0434E+00
PB-214	161.	351.92	6.9717E+00
RA-224	190.	240.99	1.2516E+01
RA-226	223.	186.10	1.2999E+01
AC-228	65.	911.21	3.5910E+00
TH-228	211.	84.37	5.3095E+01
TH-229	203.	88.47	2.4580E+00
U-232	156.	57.78	8.5930E+02
PA-233	160.	312.17	1.4584E+00
UTH-233	188.	245.34	4.5048E+02
PA-234M	37.	1001.03	1.2114E+02
TH-234	164.	63.29	3.3172E+01
U-235	229.	185.71	8.0052E-01
NP-237	205.	86.48	5.1163E+00
U-237	154.	59.54	4.4323E+00
NP-238	34.	984.45	2.7339E+00
NP-239	204.	106.12	2.1267E+00
PU-239	194.	129.30	6.4378E+03
AM-241	154.	59.54	4.2362E+00
AM-243	188.	74.67	1.3843E+00

*****
 * 222-S Laboratory Counting Room 18-NOV-1998 00:20:40.42 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T2960-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLUder

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2842.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:41.79 sec
 Dead Time: 6.7%

Verified by:

John Relyea 18 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 17-NOV-1998 21:39:31.94
 Decayed to: 17-NOV-1998 21:39:31.94
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	188.85	12884	6.48	377.75	370	22	21.3			
0	661.56*	2205003	1.49	1323.11	1315	16	0.1		CS-137	1.729E+05
0	1322.96	248	1.56	2645.99	2638	17	16.2			
~~0~~	~~1460.50*~~	~~25~~	~~2.23~~	~~2921.13~~	~~2913~~	~~18~~	~~166.5~~		~~K-40~~	~~29.3~~

JP

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2960-SAM

Acquisition date : 17-NOV-1998 21:39:31

Total number of lines in spectrum	4
Number of unidentified lines	2
Number of lines tentatively identified by NID	2 50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~2.935E+01~~	~~2.935E+01~~	~~4.886E+01~~	~~166.48~~	~~RAJ~~
CS-137	30.00Y	1.00	1.729E+05	1.729E+05	0.002E+05	0.14	
Total Activity :			1.730E+05	1.730E+05			11-19-98

Grand Total Activity :	1.730E+05	1.730E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2960-SAM

Acquisition date : 17-NOV-1998 21:39:31

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	66107.	477.59	5.8275E+02
NA-24	16.	1368.55	2.6540E+00
AR-41	25.	1293.60	4.6987E+00
SC-46	90.	1120.55	4.7448E+00
CR-51	61828.	320.08	4.0659E+02
MN-54	109.	834.83	4.1433E+00
CO-56	98.	846.76	3.9837E+00
CO-58	135.	810.78	4.5079E+00
FE-59	53.	1099.25	6.4669E+00
CO-60	32.	1332.50	3.3380E+00
ZN-65	51.	1115.55	7.1130E+00
SE-75	69936.	264.66	6.3994E+01
KR-85	34666.	514.00	1.0746E+04
SR-85	34661.	514.01	4.8608E+01
Y-88	11.	1836.06	2.7473E+00
Y-91	46.	1204.67	1.2155E+03
NB-94	101.	871.09	4.1300E+00
NB-95	245.	765.78	5.7078E+00
ZR-95	307.	756.73	1.1542E+01
ZRNB-95	491.	724.18	3.4961E+01
RU-103	41681.	497.08	5.4590E+01
RURH-106	14870.	621.93	7.2844E+02
AG-108m	499.	722.94	8.4429E+00
CD-109	58819.	88.03	1.1480E+03
AG-110M	47855.	657.76	7.1702E+01
SN-113	62796.	391.69	7.6867E+01
TE-123m	71906.	159.00	3.5853E+01
SB-124	15195.	602.73	3.6373E+01
SB-125	71388.	427.89	1.9302E+02
TE-125m	63915.	109.27	1.1608E+04
I-129	52929.	39.60	1.0984E+04
I-131	60448.	364.48	5.6018E+01
XE-131m	71884.	163.93	1.5454E+03
BA-133	60267.	356.02	7.1360E+01
CS-134	15353.	604.70	3.6716E+01
CS-136	118.	818.51	4.2494E+00
CS-138	16.	1435.86	1.2166E+01
CE-139	71698.	165.85	3.7815E+01
BA-140	25108.	537.31	1.6985E+02
LA-140	15.	1596.21	2.9901E+00
CE-141	72223.	145.44	6.2821E+01
CE-144	72737.	133.51	2.7899E+02
CEPR-144	72738.	133.51	5.5748E+02
EU-152	69735.	121.78	1.1115E+02
EU-154	36.	1274.51	9.9149E+00
EU-155	58658.	86.54	1.2996E+02
HF-181	54377.	482.18	6.8566E+01
TA-182	53721.	67.75	1.7460E+02
HG-203	67740.	279.20	4.7117E+01
BI-207	16406.	569.70	3.5965E+01
TL-208	67350.	277.36	6.0296E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2960-SAM

Acquisition date : 17-NOV-1998 21:39:31

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	52387.	46.50	7.9741E+03
BI-212	496.	727.18	1.1548E+02
PB-212	75171.	238.63	8.2817E+01
BI-214	15189.	609.31	8.0080E+01
PB-214	60598.	351.92	5.1840E+02
RA-224	74606.	240.99	9.2167E+02
RA-226	91614.	186.10	1.0034E+03
AC-228	129.	911.21	1.8067E+01
TH-228	58032.	84.37	3.5072E+03
TH-229	58842.	88.47	1.6642E+02
U-232	52479.	57.78	6.0557E+04
PA-233	63038.	312.17	1.0499E+02
UTH-233	73609.	245.34	3.3083E+04
PA-234M	63.	1001.03	5.6441E+02
TH-234	52817.	63.29	2.3164E+03
U-235	90398.	185.71	6.0628E+01
NP-237	58687.	86.48	3.4429E+02
U-237	52532.	59.54	3.1605E+02
NP-238	55.	984.45	1.2507E+01
NP-239	62943.	106.12	1.4826E+02
PU-239	72302.	129.30	4.8689E+05
AM-241	52532.	59.54	3.0197E+02
AM-243	55284.	74.67	9.3830E+01

*****

* 222-S Laboratory Counting Room 18-NOV-1998 05:45:54.49 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26738
 Sample ID: S98T2960-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

TC Wright

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2846.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:40:35.11 sec
 Dead Time: 6.6%

Verified by:

John Pely 18 NOV 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 18-NOV-1998 03:04:51.42
 Decayed to: 18-NOV-1998 03:04:51.42
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	189.32	20580	7.23	378.68	369	26 15.3			
0	661.57*	2176103	1.48	1323.13	1315	16 0.1		CS-137	1.707E+05
0	1322.86	226	2.06	2645.80	2639	14 17.3			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2960-DUP

Acquisition date : 18-NOV-1998 03:04:51

Total number of lines in spectrum	3	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	1	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.707E+05	1.707E+05	0.002E+05	0.14	
Total Activity :			1.707E+05	1.707E+05			

Grand Total Activity :	1.707E+05	1.707E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2960-DUP

Acquisition date : 18-NOV-1998 03:04:51

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	65604.	477.59	5.8053E+02
NA-24	19.	1368.55	2.8655E+00
K-40	301.	1460.75	9.6789E+01
AR-41	25.	1293.60	4.6673E+00
SC-46	74.	1120.55	4.3305E+00
CR-51	61131.	320.08	4.0430E+02
MN-54	112.	834.83	4.1932E+00
CO-56	110.	846.76	4.2000E+00
CO-58	140.	810.78	4.5885E+00
FE-59	49.	1099.25	6.2405E+00
CO-60	30.	1332.50	3.2504E+00
ZN-65	68.	1115.55	8.1694E+00
SE-75	69416.	264.66	6.3756E+01
KR-85	34169.	514.00	1.0669E+04
SR-85	34161.	514.01	4.8257E+01
Y-88	12.	1836.06	2.8399E+00
Y-91	33.	1204.67	1.0518E+03
NB-94	102.	871.09	4.1564E+00
NB-95	258.	765.78	5.8533E+00
ZR-95	305.	756.73	1.1504E+01
ZRNB-95	479.	724.18	3.4544E+01
RU-103	41170.	497.08	5.4255E+01
RURH-106	14643.	621.93	7.2288E+02
AG-108m	451.	722.94	8.0350E+00
CD-109	57804.	88.03	1.1381E+03
AG-110M	46492.	657.76	7.0676E+01
SN-113	62666.	391.69	7.6788E+01
TE-123m	70962.	159.00	3.5618E+01
SB-124	15112.	602.73	3.6273E+01
SB-125	70587.	427.89	1.9194E+02
TE-125m	63255.	109.27	1.1548E+04
I-129	52368.	39.60	1.0926E+04
I-131	60370.	364.48	5.5982E+01
XE-131m	70673.	163.93	1.5323E+03
BA-133	59765.	356.02	7.1063E+01
CS-134	15034.	604.70	3.6335E+01
CS-136	138.	818.51	4.5804E+00
CS-138	9.	1435.86	9.6397E+00
CE-139	70825.	165.85	3.7584E+01
BA-140	24565.	537.31	1.6801E+02
LA-140	10.	1596.21	2.4707E+00
CE-141	71637.	145.44	6.2566E+01
CE-144	71370.	133.51	2.7636E+02
CEPR-144	71367.	133.51	5.5221E+02
EU-152	69338.	121.78	1.1084E+02
EU-154	30.	1274.51	9.1201E+00
EU-155	57526.	86.54	1.2870E+02
HF-181	53604.	482.18	6.8078E+01
TA-182	52671.	67.75	1.7289E+02
HG-203	66435.	279.20	4.6662E+01
BI-207	16475.	569.70	3.6040E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2960-DUP

Acquisition date : 18-NOV-1998 03:04:51

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	66925.	277.36	6.0107E+02
PB-210	51821.	46.50	7.9310E+03
BI-212	482.	727.18	1.1388E+02
PB-212	73808.	238.63	8.2065E+01
BI-214	14597.	609.31	7.8512E+01
PB-214	60105.	351.92	5.1596E+02
RA-224	73638.	240.99	9.1568E+02
RA-226	90466.	186.10	9.9711E+02
AC-228	105.	911.21	1.6337E+01
TH-228	57475.	84.37	3.4904E+03
TH-229	57799.	88.47	1.6494E+02
U-232	51543.	57.78	6.0016E+04
PA-233	61927.	312.17	1.0406E+02
UTH-233	72659.	245.34	3.2870E+04
PA-234M	66.	1001.03	5.7750E+02
TH-234	52405.	63.29	2.3074E+03
U-235	89202.	185.71	6.0226E+01
NP-237	57543.	86.48	3.4093E+02
U-237	51450.	59.54	3.1279E+02
NP-238	74.	984.45	1.4360E+01
NP-239	61806.	106.12	1.4692E+02
PU-239	70904.	129.30	4.8218E+05
AM-241	51451.	59.54	2.9886E+02
AM-243	54854.	74.67	9.3465E+01

*****
 * 222-S Laboratory Counting Room 18-NOV-1998 08:48:13.06
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26738
 Sample ID: S98T002961-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Newlight

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2847.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:41:17.53 sec
 Dead Time: 7.0%

Verified by:

John Relyea 18 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 18-NOV-1998 06:06:27.61
 Decayed to: 18-NOV-1998 06:06:27.61
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	74.96*	1187	1.05	149.98	147	7	74.3		AM-243	102.
0	661.56*	2333426	1.49	1323.11	1315	16	0.1		CS-137	1.830E+05
0	1322.85	256	2.06	2645.79	2638	17	15.4			
0	1460.63*	33	2.03	2921.37	2914	15	129.3		K-40	38.2

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002961-SAM

Acquisition date : 18-NOV-1998 06:06:27

Total number of lines in spectrum	4	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	3	75.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~3.823E+01~~	~~3.823E+01~~	~~4.944E+01~~	~~129.33~~	~~82P~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~1.830E+05~~	~~1.830E+05~~	~~0.002E+05~~	~~0.13~~	
~~AM-243~~	~~7380.00Y*****~~	~~1.017E+02~~	~~1.017E+02~~	~~1.017E+02~~	~~0.756E+02~~	~~74.33~~	~~82P~~
Total Activity :			1.831E+05	1.831E+05			11-19-98

Grand Total Activity :	1.831E+05	1.831E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002961-SAM

Acquisition date : 18-NOV-1998 06:06:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	70288.	477.59	6.0086E+02
NA-24	19.	1368.55	2.8783E+00
AR-41	30.	1293.60	5.0678E+00
SC-46	84.	1120.55	4.5951E+00
CR-51	65027.	320.08	4.1696E+02
MN-54	115.	834.83	4.2450E+00
CO-56	130.	846.76	4.5628E+00
CO-58	163.	810.78	4.9263E+00
FE-59	59.	1099.25	6.7792E+00
CO-60	30.	1332.50	3.2388E+00
ZN-65	59.	1115.55	7.6596E+00
SE-75	73926.	264.66	6.5790E+01
KR-85	36431.	514.00	1.1016E+04
SR-85	36427.	514.01	4.9828E+01
Y-88	12.	1836.06	2.8565E+00
Y-91	52.	1204.67	1.2938E+03
NB-94	114.	871.09	4.3761E+00
NB-95	263.	765.78	5.9016E+00
ZR-95	346.	756.73	1.2228E+01
ZRNB-95	588.	724.18	3.8170E+01
RU-103	44626.	497.08	5.6480E+01
RURH-106	15819.	621.93	7.5123E+02
AG-108m	608.	722.94	9.2956E+00
CD-109	62006.	88.03	1.1787E+03
AG-110M	50611.	657.76	7.3732E+01
SN-113	67259.	391.69	7.9545E+01
TE-123m	76264.	159.00	3.6922E+01
SB-124	16015.	602.73	3.7337E+01
SB-125	75556.	427.89	1.9857E+02
TE-125m	67877.	109.27	1.1962E+04
I-129	56263.	39.60	1.1324E+04
I-131	64462.	364.48	5.7845E+01
XE-131m	76056.	163.93	1.5895E+03
BA-133	64210.	356.02	7.3652E+01
CS-134	16121.	604.70	3.7619E+01
CS-136	131.	818.51	4.4793E+00
CS-138	15.	1435.86	1.1931E+01
CE-139	76419.	165.85	3.9037E+01
BA-140	26559.	537.31	1.7468E+02
LA-140	14.	1596.21	2.8398E+00
CE-141	76813.	145.44	6.4783E+01
CE-144	76319.	133.51	2.8576E+02
CEPR-144	76318.	133.51	5.7100E+02
EU-152	73897.	121.78	1.1441E+02
EU-154	31.	1274.51	9.2918E+00
EU-155	61721.	86.54	1.3330E+02
HF-181	57347.	482.18	7.0409E+01
TA-182	57239.	67.75	1.8021E+02
HG-203	71849.	279.20	4.8522E+01
BI-207	17615.	569.70	3.7261E+01
TL-208	71578.	277.36	6.2156E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002961-SAM

Acquisition date : 18-NOV-1998 06:06:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	55552.	46.50	8.2110E+03
BI-212	562.	727.18	1.2280E+02
PB-212	80052.	238.63	8.5458E+01
BI-214	15884.	609.31	8.1884E+01
PB-214	64174.	351.92	5.3529E+02
RA-224	79055.	240.99	9.4874E+02
RA-226	96654.	186.10	1.0306E+03
AC-228	106.	911.21	1.6476E+01
TH-228	61475.	84.37	3.6095E+03
TH-229	62168.	88.47	1.7105E+02
U-232	55541.	57.78	6.2294E+04
PA-233	66372.	312.17	1.0772E+02
UTH-233	77679.	245.34	3.3984E+04
PA-234M	70.	1001.03	5.9326E+02
TH-234	55825.	63.29	2.3813E+03
U-235	95398.	185.71	6.2279E+01
NP-237	61729.	86.48	3.5308E+02
U-237	55847.	59.54	3.2585E+02
NP-238	65.	984.45	1.3479E+01
NP-239	66248.	106.12	1.5210E+02
PU-239	75729.	129.30	4.9828E+05
AM-241	55847.	59.54	3.1133E+02

*****

* 222-S Laboratory Counting Room 18-NOV-1998 12:08:54.65 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T002961-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

Neusig

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2848.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:41:23.21 sec
 Dead Time: 7.1%

Verified by:

RT Jones 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 09:27:03.20
 Decayed to: 18-NOV-1998 09:27:03.20
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	188.49	8011	5.02	377.02	371	15	25.3			
0	475.76	2041	2.18	951.52	949	8	45.8			
0	661.54*	2348902	1.49	1323.07	1315	16	0.1		CS-137	1.842E+05
0	1322.75	265	1.95	2645.57	2638	19	17.1			
0	1460.46*	49	2.09	2921.04	2911	18	83.1		K-40	57.0

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002961-DUP

Acquisition date : 18-NOV-1998 09:27:03

Total number of lines in spectrum 5
 Number of unidentified lines 3
 Number of lines tentatively identified by NID 2 40.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~5.699E+01~~	~~5.699E+01~~	~~4.737E+01~~	~~83.13~~	~~K~~
CS-137	30.00Y	1.00	1.842E+05	1.842E+05	0.002E+05	0.13	
Total Activity :			1.843E+05	1.843E+05			11-19-98

Grand Total Activity : 1.843E+05 1.843E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
 Sample ID : S98T002961-DUP

Page : 3
 Acquisition date : 18-NOV-1998 09:27:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	70831.	477.59	6.0317E+02
NA-24	18.	1368.55	2.8405E+00
AR-41	35.	1293.60	5.4927E+00
SC-46	78.	1120.55	4.4368E+00
CR-51	65696.	320.08	4.1909E+02
MN-54	125.	834.83	4.4138E+00
CO-56	115.	846.76	4.2890E+00
CO-58	148.	810.78	4.6993E+00
FE-59	59.	1099.25	6.7986E+00
CO-60	45.	1332.50	3.9151E+00
ZN-65	49.	1115.55	7.0189E+00
SE-75	74804.	264.66	6.6179E+01
KR-85	36885.	514.00	1.1084E+04
SR-85	36879.	514.01	5.0135E+01
Y-88	4.	1836.06	1.8058E+00
Y-91	44.	1204.67	1.1908E+03
NB-94	92.	871.09	3.9544E+00
NB-95	283.	765.78	6.1217E+00
ZR-95	334.	756.73	1.2026E+01
ZRNB-95	642.	724.18	3.9836E+01
RU-103	44819.	497.08	5.6602E+01
RURH-106	15630.	621.93	7.4674E+02
AG-108m	634.	722.94	9.4879E+00
CD-109	62537.	88.03	1.1837E+03
AG-110M	52977.	657.76	7.5431E+01
SN-113	67381.	391.69	7.9617E+01
TE-123m	76554.	159.00	3.6992E+01
SB-124	16155.	602.73	3.7499E+01
SB-125	75977.	427.89	1.9912E+02
TE-125m	68412.	109.27	1.2008E+04
I-129	56724.	39.60	1.1370E+04
I-131	65109.	364.48	5.8134E+01
XE-131m	76461.	163.93	1.5937E+03
BA-133	64678.	356.02	7.3919E+01
CS-134	15973.	604.70	3.7447E+01
CS-136	125.	818.51	4.3661E+00
CS-138	14.	1435.86	1.1641E+01
CE-139	76545.	165.85	3.9069E+01
BA-140	26735.	537.31	1.7525E+02
LA-140	16.	1596.21	3.0156E+00
CE-141	77440.	145.44	6.5046E+01
CE-144	77291.	133.51	2.8757E+02
CEPR-144	77292.	133.51	5.7463E+02
EU-152	74342.	121.78	1.1476E+02
EU-154	26.	1274.51	8.5577E+00
EU-155	62346.	86.54	1.3397E+02
HF-181	57838.	482.18	7.0709E+01
TA-182	56988.	67.75	1.7982E+02
HG-203	72221.	279.20	4.8647E+01
BI-207	17825.	569.70	3.7481E+01
TL-208	72042.	277.36	6.2357E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002961-DUP

Acquisition date : 18-NOV-1998 09:27:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	55722.	46.50	8.2235E+03
BI-212	603.	727.18	1.2705E+02
PB-212	79719.	238.63	8.5280E+01
BI-214	16283.	609.31	8.2900E+01
PB-214	64613.	351.92	5.3741E+02
RA-224	79048.	240.99	9.4870E+02
RA-226	97772.	186.10	1.0365E+03
AC-228	122.	911.21	1.7588E+01
TH-228	61801.	84.37	3.6190E+03
TH-229	62589.	88.47	1.7162E+02
U-232	56261.	57.78	6.2696E+04
PA-233	66890.	312.17	1.0814E+02
UTH-233	78294.	245.34	3.4118E+04
PA-234M	67.	1001.03	5.8359E+02
TH-234	56541.	63.29	2.3964E+03
U-235	96553.	185.71	6.2654E+01
NP-237	62384.	86.48	3.5495E+02
U-237	56426.	59.54	3.2753E+02
NP-238	53.	984.45	1.2305E+01
NP-239	66762.	106.12	1.5269E+02
PU-239	76029.	129.30	4.9926E+05
AM-241	56426.	59.54	3.1294E+02
AM-243	59595.	74.67	9.7410E+01

*****
 * 222-S Laboratory Counting Room 19-NOV-1998 10:50:31.96
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T002962-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

NE Wright

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: DKA300:[SPEC.GEA2]2G2849.CNF
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:42:35.57 sec
 Dead Time: 7.7%

Verified by:

RA Jones 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 12:50:02.31
 Decayed to: 18-NOV-1998 12:50:02.31
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.54*	2612413	1.49	1323.07	1315	16	0.1		CS-137	2.049E+05
2	1322.68	301	2.20	2645.43	2637	33	12.8	1.80E+00		
2	1331.98	30	2.20	2664.04	2637	33	55.4			
0	1460.65*	38	1.84	2921.43	2913	16115.7			K-40	44.2

Summary of Nuclide Activity
Sample ID : S98T002962-SAM

Page : 2
Acquisition date : 18-NOV-1998 12:50:02

Total number of lines in spectrum 4
Number of unidentified lines 1
Number of lines tentatively identified by NID 3 75.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~4.421E+01~~	~~4.421E+01~~	~~5.115E+01~~	~~115.69~~	~~NA~~
CS-137	30.00Y	1.00	2.049E+05	2.049E+05	0.003E+05	0.13	

Total Activity : 2.049E+05 2.049E+05

11-19-98

Grand Total Activity : 2.049E+05 2.049E+05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002962-SAM

Acquisition date : 18-NOV-1998 12:50:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	78847.	477.59	6.3631E+02
NA-24	20.	1368.55	2.9721E+00
AR-41	30.	1293.60	5.0981E+00
SC-46	103.	1120.55	5.0546E+00
CR-51	73329.	320.08	4.4272E+02
MN-54	154.	834.83	4.8860E+00
CO-56	123.	846.76	4.4407E+00
CO-58	149.	810.78	4.7273E+00
FE-59	64.	1099.25	7.0641E+00
CO-60	34.	1332.50	3.4379E+00
ZN-65	76.	1115.55	8.5834E+00
SE-75	83092.	264.66	6.9741E+01
KR-85	40927.	514.00	1.1674E+04
SR-85	40924.	514.01	5.2805E+01
Y-88	8.	1836.06	2.3844E+00
Y-91	55.	1204.67	1.3255E+03
NB-94	101.	871.09	4.1373E+00
NB-95	344.	765.78	6.7227E+00
ZR-95	385.	756.73	1.2875E+01
ZRNB-95	707.	724.18	4.1781E+01
RU-103	50103.	497.08	5.9837E+01
RURH-106	17847.	621.93	7.9771E+02
AG-108m	717.	722.94	1.0078E+01
CD-109	70016.	88.03	1.2523E+03
AG-110M	58945.	657.76	7.9556E+01
SN-113	75128.	391.69	8.4060E+01
TE-123m	85291.	159.00	3.9041E+01
SB-124	17887.	602.73	3.9449E+01
SB-125	84922.	427.89	2.1049E+02
TE-125m	76064.	109.27	1.2661E+04
I-129	63041.	39.60	1.1985E+04
I-131	72208.	364.48	6.1217E+01
XE-131m	85049.	163.93	1.6807E+03
BA-133	71486.	356.02	7.7704E+01
CS-134	17873.	604.70	3.9602E+01
CS-136	141.	818.51	4.6271E+00
CS-138	21.	1435.86	1.3841E+01
CE-139	85489.	165.85	4.1284E+01
BA-140	29847.	537.31	1.8514E+02
LA-140	11.	1596.21	2.5875E+00
CE-141	86445.	145.44	6.8717E+01
CE-144	85391.	133.51	3.0223E+02
CEPR-144	85391.	133.51	6.0392E+02
EU-152	82420.	121.78	1.2082E+02
EU-154	36.	1274.51	9.9073E+00
EU-155	69502.	86.54	1.4143E+02
HF-181	64434.	482.18	7.4623E+01
TA-182	63684.	67.75	1.9006E+02
HG-203	80576.	279.20	5.1378E+01
BI-207	19877.	569.70	3.9570E+01
TL-208	80345.	277.36	6.5845E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002962-SAM

Acquisition date : 18-NOV-1998 12:50:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	61614.	46.50	8.6467E+03
BI-212	690.	727.18	1.3565E+02
PB-212	89323.	238.63	9.0261E+01
BI-214	17964.	609.31	8.7055E+01
PB-214	72196.	351.92	5.7196E+02
RA-224	88375.	240.99	1.0031E+03
RA-226	108294.	186.10	1.0908E+03
AC-228	146.	911.21	1.9145E+01
TH-228	69158.	84.37	3.8279E+03
TH-229	70105.	88.47	1.8161E+02
U-232	62312.	57.78	6.5973E+04
PA-233	74597.	312.17	1.1419E+02
UTH-233	87324.	245.34	3.6028E+04
PA-234M	74.	1001.03	6.0970E+02
TH-234	62883.	63.29	2.5270E+03
U-235	106985.	185.71	6.5946E+01
NP-237	69496.	86.48	3.7459E+02
U-237	61959.	59.54	3.4320E+02
NP-238	83.	984.45	1.5188E+01
NP-239	74078.	106.12	1.6084E+02
PU-239	85267.	129.30	5.2866E+05
AM-241	61959.	59.54	3.2789E+02
AM-243	65919.	74.67	1.0244E+02

*****

* 222-S Laboratory Counting Room 18-NOV-1998 18:42:25.13 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T002962-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

gi Kabe

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2851.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:42:34.76 sec
 Dead Time: 7.7%

Verified by:

RA Jones 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 15:59:20.95
 Decayed to: 18-NOV-1998 15:59:20.95
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	27.67	11855	1.78	55.41	54	6	7.0			
0	661.55*	2608649	1.49	1323.08	1315	16	0.1		CS-137	2.046E+05
0	1322.76	323	1.91	2645.60	2636	18	14.5			
0	1460.59*	64	1.59	2921.31	2913	17	66.8		K-40	74.0

Summary of Nuclide Activity
Sample ID : S98T002962-DUP

Page : 2
Acquisition date : 18-NOV-1998 15:59:20

Total number of lines in spectrum 4
Number of unidentified lines 2
Number of lines tentatively identified by NID 2 50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~7.403E+01~~	~~7.403E+01~~	~~4.944E+01~~	~~66.78~~	~~1219~~
CS-137	30.00Y	1.00	2.046E+05	2.046E+05	0.003E+05	0.13	
Total Activity :			2.047E+05	2.047E+05			11-19-98

Grand Total Activity : 2.047E+05 2.047E+05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002962-DUP

Acquisition date : 18-NOV-1998 15:59:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	78622.	477.59	6.3541E+02
NA-24	23.	1368.55	3.1232E+00
AR-41	28.	1293.60	4.9764E+00
SC-46	86.	1120.55	4.6382E+00
CR-51	72983.	320.08	4.4168E+02
MN-54	148.	834.83	4.7815E+00
CO-56	116.	846.76	4.3093E+00
CO-58	175.	810.78	5.0912E+00
FE-59	60.	1099.25	6.8452E+00
CO-60	28.	1332.50	3.1361E+00
ZN-65	64.	1115.55	7.9315E+00
SE-75	83229.	264.66	6.9799E+01
KR-85	41038.	514.00	1.1689E+04
SR-85	41030.	514.01	5.2873E+01
Y-88	7.	1836.06	2.3130E+00
Y-91	52.	1204.67	1.2938E+03
NB-94	125.	871.09	4.5746E+00
NB-95	317.	765.78	6.4667E+00
ZR-95	373.	756.73	1.2684E+01
ZRNB-95	739.	724.18	4.2691E+01
RU-103	49961.	497.08	5.9752E+01
RURH-106	17333.	621.93	7.8618E+02
AG-108m	754.	722.94	1.0331E+01
CD-109	69250.	88.03	1.2454E+03
AG-110M	58418.	657.76	7.9201E+01
SN-113	74812.	391.69	8.3884E+01
TE-123m	84765.	159.00	3.8921E+01
SB-124	18127.	602.73	3.9711E+01
SB-125	84554.	427.89	2.1003E+02
TE-125m	75804.	109.27	1.2639E+04
I-129	63177.	39.60	1.1998E+04
I-131	72261.	364.48	6.1239E+01
XE-131m	85667.	163.93	1.6868E+03
BA-133	71788.	356.02	7.7867E+01
CS-134	18059.	604.70	3.9807E+01
CS-136	149.	818.51	4.7483E+00
CS-138	21.	1435.86	1.3822E+01
CE-139	85313.	165.85	4.1242E+01
BA-140	29974.	537.31	1.8553E+02
LA-140	10.	1596.21	2.4956E+00
CE-141	85824.	145.44	6.8470E+01
CE-144	85497.	133.51	3.0242E+02
CEPR-144	85501.	133.51	6.0431E+02
EU-152	82592.	121.78	1.2094E+02
EU-154	23.	1274.51	8.1315E+00
EU-155	69152.	86.54	1.4107E+02
HF-181	64379.	482.18	7.4591E+01
TA-182	63107.	67.75	1.8920E+02
HG-203	79811.	279.20	5.1134E+01
BI-207	19767.	569.70	3.9461E+01
TL-208	79694.	277.36	6.5578E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002962-DUP

Acquisition date : 18-NOV-1998 15:59:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	61685.	46.50	8.6517E+03
BI-212	703.	727.18	1.3689E+02
PB-212	88785.	238.63	8.9990E+01
BI-214	17932.	609.31	8.6979E+01
PB-214	72159.	351.92	5.7177E+02
RA-224	88962.	240.99	1.0064E+03
RA-226	108007.	186.10	1.0893E+03
AC-228	128.	911.21	1.8003E+01
TH-228	69140.	84.37	3.8274E+03
TH-229	69266.	88.47	1.8053E+02
U-232	62039.	57.78	6.5829E+04
PA-233	74100.	312.17	1.1381E+02
UTH-233	87228.	245.34	3.6008E+04
PA-234M	71.	1001.03	5.9757E+02
TH-234	62410.	63.29	2.5175E+03
U-235	106573.	185.71	6.5819E+01
NP-237	69171.	86.48	3.7371E+02
U-237	62274.	59.54	3.4406E+02
NP-238	74.	984.45	1.4383E+01
NP-239	73923.	106.12	1.6068E+02
PU-239	85039.	129.30	5.2795E+05
AM-241	62274.	59.54	3.2872E+02
AM-243	65852.	74.67	1.0239E+02

*****
 * 222-S Laboratory Counting Room 19-NOV-1998 00:04:43.10
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T2963-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2853.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:44:30.02 sec
 Dead Time: 8.8%

Verified by:

RA Jones 11-15-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 21:19:45.72
 Decayed to: 18-NOV-1998 21:19:45.72
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.61*	3011248	1.50	1323.21	1315	16	0.1		CS-137	2.362E+05
0	1322.72	455	2.27	2645.52	2637	17	10.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2963-SAM

Acquisition date : 18-NOV-1998 21:19:45

Total number of lines in spectrum	2	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	1	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.362E+05	2.362E+05	0.003E+05	0.12	
Total Activity :			2.362E+05	2.362E+05			

Grand Total Activity :	2.362E+05	2.362E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2963-SAM

Acquisition date : 18-NOV-1998 21:19:45

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	91207.	477.59	6.8428E+02
NA-24	22.	1368.55	3.0638E+00
K-40	332.	1460.75	1.0150E+02
AR-41	26.	1293.60	4.7863E+00
SC-46	92.	1120.55	4.7818E+00
CR-51	85183.	320.08	4.7710E+02
MN-54	187.	834.83	5.3527E+00
CO-56	164.	846.76	5.0835E+00
CO-58	217.	810.78	5.6461E+00
FE-59	69.	1099.25	7.2915E+00
CO-60	33.	1332.50	3.3915E+00
ZN-65	77.	1115.55	8.6523E+00
SE-75	96455.	264.66	7.5130E+01
KR-85	47486.	514.00	1.2572E+04
SR-85	47482.	514.01	5.6867E+01
Y-88	11.	1836.06	2.7420E+00
Y-91	41.	1204.67	1.1552E+03
NB-94	169.	871.09	5.2837E+00
NB-95	451.	765.78	7.6746E+00
ZR-95	545.	756.73	1.5255E+01
ZRNB-95	908.	724.18	4.7227E+01
RU-103	57743.	497.08	6.4226E+01
RURH-106	20116.	621.93	8.4668E+02
AG-108m	914.	722.94	1.1348E+01
CD-109	80075.	88.03	1.3391E+03
AG-110M	61201.	657.76	8.1061E+01
SN-113	86264.	391.69	9.0063E+01
TE-123m	98340.	159.00	4.1916E+01
SB-124	20844.	602.73	4.2572E+01
SB-125	97801.	427.89	2.2586E+02
TE-125m	87744.	109.27	1.3596E+04
I-129	72893.	39.60	1.2886E+04
I-131	82885.	364.48	6.5581E+01
XE-131m	98621.	163.93	1.8097E+03
BA-133	83245.	356.02	8.3838E+01
CS-134	20871.	604.70	4.2780E+01
CS-136	200.	818.51	5.4763E+00
CS-138	13.	1435.86	1.1348E+01
CE-139	98297.	165.85	4.4263E+01
BA-140	34881.	537.31	2.0010E+02
LA-140	13.	1596.21	2.7523E+00
CE-141	98733.	145.44	7.3430E+01
CE-144	98993.	133.51	3.2537E+02
CEPR-144	98992.	133.51	6.5015E+02
EU-152	95842.	121.78	1.3027E+02
EU-154	32.	1274.51	9.4384E+00
EU-155	79986.	86.54	1.5170E+02
HF-181	74422.	482.18	8.0187E+01
TA-182	73660.	67.75	2.0438E+02
HG-203	92170.	279.20	5.4944E+01
BI-207	23159.	569.70	4.2700E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2963-SAM

Acquisition date : 18-NOV-1998 21:19:45

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	93048.	277.36	7.0849E+02
PB-210	72521.	46.50	9.3796E+03
BI-212	923.	727.18	1.5650E+02
PB-212	103066.	238.63	9.6944E+01
BI-214	20652.	609.31	9.3316E+01
PB-214	83145.	351.92	6.2046E+02
RA-224	102361.	240.99	1.0795E+03
RA-226	124889.	186.10	1.1712E+03
AC-228	162.	911.21	2.0104E+01
TH-228	80075.	84.37	4.1183E+03
TH-229	80197.	88.47	1.9422E+02
U-232	72073.	57.78	7.0941E+04
PA-233	85932.	312.17	1.2254E+02
UTH-233	100705.	245.34	3.8684E+04
PA-234M	93.	1001.03	6.7806E+02
TH-234	72537.	63.29	2.7136E+03
U-235	123279.	185.71	7.0781E+01
NP-237	80035.	86.48	4.0193E+02
U-237	72186.	59.54	3.7040E+02
NP-238	100.	984.45	1.6539E+01
NP-239	85916.	106.12	1.7323E+02
PU-239	97805.	129.30	5.6612E+05
AM-241	72186.	59.54	3.5385E+02
AM-243	76531.	74.67	1.1036E+02

*****

* 222-S Laboratory Counting Room 19-NOV-1998 03:35:57.63 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T2963-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLUaker

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2855.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:44:38.78 sec
 Dead Time: 8.9%

Verified by:

RAYmer 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 00:50:50.92
 Decayed to: 19-NOV-1998 00:50:50.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	188.71	11872	5.07	377.46	371	16 20.5			
0	661.63*	3036345	1.50	1323.25	1315	16 0.1		CS-137	2.381E+05
0	1322.79	475	1.80	2645.66	2638	15 10.6			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2963-DUP

Acquisition date : 19-NOV-1998 00:50:50

Total number of lines in spectrum	3	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	1	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.381E+05	2.381E+05	0.003E+05	0.12	
Total Activity :			2.381E+05	2.381E+05			

Grand Total Activity :	2.381E+05	2.381E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2963-DUP

Acquisition date : 19-NOV-1998 00:50:50

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	92755.	477.59	6.9005E+02
NA-24	19.	1368.55	2.8940E+00
K-40	312.	1460.75	9.8561E+01
AR-41	31.	1293.60	5.2320E+00
SC-46	116.	1120.55	5.3389E+00
CR-51	85222.	320.08	4.7721E+02
MN-54	189.	834.83	5.3822E+00
CO-56	164.	846.76	5.0839E+00
CO-58	190.	810.78	5.3071E+00
FE-59	86.	1099.25	8.0754E+00
CO-60	34.	1332.50	3.4371E+00
ZN-65	81.	1115.55	8.8715E+00
SE-75	96976.	264.66	7.5332E+01
KR-85	47964.	514.00	1.2635E+04
SR-85	47958.	514.01	5.7151E+01
Y-88	8.	1836.06	2.3985E+00
Y-91	47.	1204.67	1.2356E+03
NB-94	134.	871.09	4.7193E+00
NB-95	438.	765.78	7.5625E+00
ZR-95	555.	756.73	1.5382E+01
ZRNB-95	937.	724.18	4.7958E+01
RU-103	58005.	497.08	6.4372E+01
RURH-106	20564.	621.93	8.5603E+02
AG-108m	952.	722.94	1.1581E+01
CD-109	80972.	88.03	1.3465E+03
AG-110M	59464.	657.76	7.9905E+01
SN-113	87406.	391.69	9.0655E+01
TE-123m	99451.	159.00	4.2152E+01
SB-124	21215.	602.73	4.2947E+01
SB-125	99542.	427.89	2.2785E+02
TE-125m	88041.	109.27	1.3619E+04
I-129	73932.	39.60	1.2977E+04
I-131	84293.	364.48	6.6135E+01
XE-131m	99773.	163.93	1.8202E+03
BA-133	84061.	356.02	8.4248E+01
CS-134	20996.	604.70	4.2908E+01
CS-136	195.	818.51	5.4060E+00
CS-138	20.	1435.86	1.3754E+01
CE-139	99834.	165.85	4.4607E+01
BA-140	35024.	537.31	2.0051E+02
LA-140	16.	1596.21	3.0061E+00
CE-141	99739.	145.44	7.3803E+01
CE-144	99898.	133.51	3.2685E+02
CEPR-144	99894.	133.51	6.5310E+02
EU-152	96548.	121.78	1.3074E+02
EU-154	27.	1274.51	8.7671E+00
EU-155	80913.	86.54	1.5258E+02
HF-181	75542.	482.18	8.0787E+01
TA-182	73910.	67.75	2.0472E+02
HG-203	93724.	279.20	5.5404E+01
BI-207	23308.	569.70	4.2837E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2963-DUP

Acquisition date : 19-NOV-1998 00:50:50

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	93487.	277.36	7.1016E+02
PB-210	72343.	46.50	9.3682E+03
BI-212	888.	727.18	1.5353E+02
PB-212	104340.	238.63	9.7540E+01
BI-214	20660.	609.31	9.3333E+01
PB-214	84118.	351.92	6.2459E+02
RA-224	103364.	240.99	1.0848E+03
RA-226	126361.	186.10	1.1781E+03
AC-228	175.	911.21	2.0850E+01
TH-228	80503.	84.37	4.1293E+03
TH-229	81018.	88.47	1.9521E+02
U-232	72443.	57.78	7.1122E+04
PA-233	86720.	312.17	1.2310E+02
UTH-233	101254.	245.34	3.8790E+04
PA-234M	82.	1001.03	6.4039E+02
TH-234	73473.	63.29	2.7310E+03
U-235	124752.	185.71	7.1202E+01
NP-237	80947.	86.48	4.0421E+02
U-237	72419.	59.54	3.7100E+02
NP-238	93.	984.45	1.6002E+01
NP-239	86646.	106.12	1.7396E+02
PU-239	98982.	129.30	5.6951E+05
AM-241	72419.	59.54	3.5442E+02
AM-243	77189.	74.67	1.1083E+02

*****
 * 222-S Laboratory Counting Room 19-NOV-1998 06:23:03.15 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T2964-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

newlight

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2856.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:43:39.36 sec
 Dead Time: 8.3%

Verified by:

RA Jones 11-19-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 03:38:57.57
 Decayed to: 19-NOV-1998 03:38:57.57
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	75.02*	1195	0.96	150.11	148	6 72.1		AM-243	102.
0	661.62*	2833947	1.49	1323.23	1315	16 0.1		CS-137	2.223E+05
0	1323.06	375	1.80	2646.20	2637	19 13.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2964-SAM

Acquisition date : 19-NOV-1998 03:38:57

Total number of lines in spectrum	3	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	2	66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.223E+05	2.223E+05	0.003E+05	0.12	
~~AM-243~~	~~7380.00Y*****~~		~~1.023E+02~~	~~1.023E+02~~	~~0.737E+02~~	~~72.06~~	~~PAJ~~
Total Activity :			2.224E+05	2.224E+05			

11-19-98

Grand Total Activity :	2.224E+05	2.224E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2964-SAM

Acquisition date : 19-NOV-1998 03:38:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	85399.	477.59	6.6217E+02
NA-24	17.	1368.55	2.7325E+00
K-40	326.	1460.75	1.0060E+02
AR-41	32.	1293.60	5.3213E+00
SC-46	100.	1120.55	4.9773E+00
CR-51	80236.	320.08	4.6306E+02
MN-54	142.	834.83	4.6902E+00
CO-56	138.	846.76	4.6938E+00
CO-58	217.	810.78	5.6482E+00
FE-59	75.	1099.25	7.5698E+00
CO-60	25.	1332.50	2.9752E+00
ZN-65	83.	1115.55	8.9598E+00
SE-75	90818.	264.66	7.2905E+01
KR-85	44345.	514.00	1.2150E+04
SR-85	44339.	514.01	5.4958E+01
Y-88	15.	1836.06	3.1230E+00
Y-91	51.	1204.67	1.2799E+03
NB-94	104.	871.09	4.1913E+00
NB-95	382.	765.78	7.0776E+00
ZR-95	513.	756.73	1.4812E+01
ZRNB-95	818.	724.18	4.4862E+01
RU-103	54024.	497.08	6.2128E+01
RURH-106	19061.	621.93	8.2427E+02
AG-108m	816.	722.94	1.0738E+01
CD-109	74874.	88.03	1.2949E+03
AG-110M	55836.	657.76	7.7435E+01
SN-113	81782.	391.69	8.7696E+01
TE-123m	93063.	159.00	4.0778E+01
SB-124	19296.	602.73	4.0967E+01
SB-125	91919.	427.89	2.1897E+02
TE-125m	82209.	109.27	1.3161E+04
I-129	68441.	39.60	1.2487E+04
I-131	78420.	364.48	6.3792E+01
XE-131m	92689.	163.93	1.7545E+03
BA-133	77819.	356.02	8.1066E+01
CS-134	19185.	604.70	4.1024E+01
CS-136	170.	818.51	5.0635E+00
CS-138	20.	1435.86	1.3559E+01
CE-139	92878.	165.85	4.3028E+01
BA-140	32319.	537.31	1.9264E+02
LA-140	13.	1596.21	2.7890E+00
CE-141	93446.	145.44	7.1440E+01
CE-144	92856.	133.51	3.1514E+02
CEPR-144	92853.	133.51	6.2971E+02
EU-152	89611.	121.78	1.2597E+02
EU-154	32.	1274.51	9.3764E+00
EU-155	75153.	86.54	1.4706E+02
HF-181	69932.	482.18	7.7735E+01
TA-182	68985.	67.75	1.9780E+02
HG-203	86440.	279.20	5.3211E+01
BI-207	21664.	569.70	4.1304E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2964-SAM

Acquisition date : 19-NOV-1998 03:38:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	86799.	277.36	6.8433E+02
PB-210	67627.	46.50	9.0581E+03
BI-212	769.	727.18	1.4312E+02
PB-212	96893.	238.63	9.4001E+01
BI-214	19508.	609.31	9.0705E+01
PB-214	77858.	351.92	5.9756E+02
RA-224	96137.	240.99	1.0462E+03
RA-226	117110.	186.10	1.1342E+03
AC-228	129.	911.21	1.8036E+01
TH-228	75076.	84.37	3.9880E+03
TH-229	74932.	88.47	1.8775E+02
U-232	67317.	57.78	6.8565E+04
PA-233	81498.	312.17	1.1935E+02
UTH-233	95269.	245.34	3.7628E+04
PA-234M	69.	1001.03	5.9265E+02
TH-234	68165.	63.29	2.6307E+03
U-235	115588.	185.71	6.8541E+01
NP-237	75203.	86.48	3.8963E+02
U-237	67837.	59.54	3.5908E+02
NP-238	78.	984.45	1.4750E+01
NP-239	80811.	106.12	1.6800E+02
PU-239	92501.	129.30	5.5059E+05
AM-241	67837.	59.54	3.4305E+02

*****

* 222-S Laboratory Counting Room 19-NOV-1998 09:26:33.03 *

*****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26738
 Sample ID: S98T002964-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

REU J. H. J.

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2857.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:43:41.08 sec
 Dead Time: 8.4%

Verified by:

REU J. H. J. 11-19-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 06:42:22.79
 Decayed to: 19-NOV-1998 06:42:22.79
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:09:16.15

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	189.97	22540	7.40	379.97	370	26 16.1			
0	661.61*	2834711	1.49	1323.21	1315	16 0.1		CS-137	2.223E+05
0	913.09	100	4.51	1826.18	1817	21 67.6			
0	1322.88	370	1.72	2645.83	2639	17 12.7			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002964-DUP

Acquisition date : 19-NOV-1998 06:42:22

Total number of lines in spectrum	4	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	1	25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.223E+05	2.223E+05	0.003E+05	0.12	
Total Activity :			2.223E+05	2.223E+05			

Grand Total Activity :	2.223E+05	2.223E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002964-DUP

Acquisition date : 19-NOV-1998 06:42:22

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	86186.	477.59	6.6521E+02
NA-24	12.	1368.55	2.3769E+00
K-40	303.	1460.75	9.7181E+01
AR-41	28.	1293.60	5.0149E+00
SC-46	83.	1120.55	4.5646E+00
CR-51	79892.	320.08	4.6207E+02
MN-54	156.	834.83	4.9121E+00
CO-56	155.	846.76	4.9531E+00
CO-58	203.	810.78	5.4679E+00
FE-59	65.	1099.25	7.0767E+00
CO-60	32.	1332.50	3.3548E+00
ZN-65	81.	1115.55	8.8481E+00
SE-75	90402.	264.66	7.2739E+01
KR-85	44227.	514.00	1.2134E+04
SR-85	44222.	514.01	5.4885E+01
Y-88	9.	1836.06	2.5192E+00
Y-91	43.	1204.67	1.1886E+03
NB-94	139.	871.09	4.8112E+00
NB-95	435.	765.78	7.5373E+00
ZR-95	433.	756.73	1.3632E+01
ZRNB-95	843.	724.18	4.5526E+01
RU-103	54782.	497.08	6.2562E+01
RURH-106	19394.	621.93	8.3142E+02
AG-108m	850.	722.94	1.0956E+01
CD-109	75484.	88.03	1.3002E+03
AG-110M	56870.	657.76	7.8147E+01
SN-113	81834.	391.69	8.7724E+01
TE-123m	92655.	159.00	4.0689E+01
SB-124	19362.	602.73	4.1037E+01
SB-125	92118.	427.89	2.1921E+02
TE-125m	83000.	109.27	1.3224E+04
I-129	68804.	39.60	1.2520E+04
I-131	78807.	364.48	6.3949E+01
XE-131m	92482.	163.93	1.7525E+03
BA-133	78159.	356.02	8.1242E+01
CS-134	19361.	604.70	4.1210E+01
CS-136	178.	818.51	5.1815E+00
CS-138	15.	1435.86	1.1949E+01
CE-139	92520.	165.85	4.2945E+01
BA-140	32567.	537.31	1.9337E+02
LA-140	14.	1596.21	2.8940E+00
CE-141	92899.	145.44	7.1231E+01
CE-144	93058.	133.51	3.1548E+02
CEPR-144	93060.	133.51	6.3041E+02
EU-152	89710.	121.78	1.2604E+02
EU-154	35.	1274.51	9.7353E+00
EU-155	75456.	86.54	1.4735E+02
HF-181	69814.	482.18	7.7670E+01
TA-182	69327.	67.75	1.9829E+02
HG-203	87155.	279.20	5.3430E+01
BI-207	21518.	569.70	4.1165E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002964-DUP

Acquisition date : 19-NOV-1998 06:42:22

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	87405.	277.36	6.8671E+02
PB-210	67643.	46.50	9.0592E+03
BI-212	816.	727.18	1.4729E+02
PB-212	97372.	238.63	9.4233E+01
BI-214	19508.	609.31	9.0704E+01
PB-214	78496.	351.92	6.0010E+02
RA-224	96573.	240.99	1.0486E+03
RA-226	117428.	186.10	1.1358E+03
AC-228	148.	911.21	1.9270E+01
TH-228	74946.	84.37	3.9845E+03
TH-229	75492.	88.47	1.8845E+02
U-232	67513.	57.78	6.8665E+04
PA-233	81084.	312.17	1.1904E+02
UTH-233	95176.	245.34	3.7609E+04
PA-234M	85.	1001.03	6.5114E+02
TH-234	68383.	63.29	2.6349E+03
U-235	115780.	185.71	6.8598E+01
NP-237	75481.	86.48	3.9035E+02
U-237	67623.	59.54	3.5852E+02
NP-238	84.	984.45	1.5208E+01
NP-239	80880.	106.12	1.6807E+02
PU-239	92332.	129.30	5.5008E+05
AM-241	67623.	59.54	3.4251E+02
AM-243	71579.	74.67	1.0673E+02

LABCORE Completed Worklist Report for Worklist# 26841

Analyst: gl

Instrument: GEA01

Book#: 28B57

Method: LA-548-121

Rev/Mod F-0

Worklist Comment: xBY-105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		GEA-01 C060-02	LIQUID	4.512e-03	4.81e-03	106.605 %	Recovery
1 STD	0		GEA-01 C060-02R	LIQUID	1	3.14	3.140 %	Ct Error
1 STD	0		GEA-01 CS13702	LIQUID	7.448e-03	7.58e-03	101.772 %	Recovery
1 STD	0		GEA-01 CS13702R	LIQUID	1	2.86	2.860 %	Ct Error
2 BLNK	0		GEA-01 C060-02	LIQUID	1	<2.71e-3		uCi/mL
2 BLNK	0		GEA-01 CS13702	LIQUID	1	8.51e-03	0.009	uCi/mL
3 SAMPLE	S98T003006	0	GEA-01 C060-02	LIQUID	N/A	< 2.898e-03	2.90e-003	uCi/mL
3 SAMPLE	S98T003006	0	GEA-01 C060-02R	LIQUID	N/A	n/a		% Ct. Error
3 SAMPLE	S98T003006	0	GEA-01 CS13702	LIQUID	N/A	2.100e+02		uCi/mL
3 SAMPLE	S98T003006	0	GEA-01 CS13702R	LIQUID	N/A	0.0900		% Ct. Error
4 DUP	S98T003006	0	GEA-01 C060-02	LIQUID	<2.90e-3	<3.05e-3		RFD
4 DUP	S98T003006	0	GEA-01 C060-02R	LIQUID	1	n/a		% Ct. Error
4 DUP	S98T003006	0	GEA-01 CS13702	LIQUID	2.10e+02	2.10e+02	0.000	RFD
4 DUP	S98T003006	0	GEA-01 CS13702R	LIQUID	1	0.0900	0.090	% Ct. Error
5 SAMPLE	S98T003007	0	GEA-01 C060-02	LIQUID	N/A	< 3.323e-03	3.32e-003	uCi/mL
5 SAMPLE	S98T003007	0	GEA-01 C060-02R	LIQUID	N/A	n/a		% Ct. Error
5 SAMPLE	S98T003007	0	GEA-01 CS13702	LIQUID	N/A	1.980e+02		uCi/mL
5 SAMPLE	S98T003007	0	GEA-01 CS13702R	LIQUID	N/A	0.0900		% Ct. Error
6 DUP	S98T003007	0	GEA-01 C060-02	LIQUID	<3.32e-3	<2.86e-3		RFD
6 DUP	S98T003007	0	GEA-01 C060-02R	LIQUID	1	n/a		% Ct. Error
6 DUP	S98T003007	0	GEA-01 CS13702	LIQUID	1.98e+02	2.00e+02	1.005	RFD
6 DUP	S98T003007	0	GEA-01 CS13702R	LIQUID	1	0.0900	0.090	% Ct. Error
7 SAMPLE	S98T003008	0	GEA-01 C060-02	LIQUID	N/A	< 3.047e-03	3.05e-003	uCi/mL
7 SAMPLE	S98T003008	0	GEA-01 C060-02R	LIQUID	N/A	n/a		% Ct. Error
7 SAMPLE	S98T003008	0	GEA-01 CS13702	LIQUID	N/A	2.150e+02		uCi/mL
7 SAMPLE	S98T003008	0	GEA-01 CS13702R	LIQUID	N/A	0.0900		% Ct. Error
8 DUP	S98T003008	0	GEA-01 C060-02	LIQUID	<3.05e-3	<3.14e-3		RFD
8 DUP	S98T003008	0	GEA-01 C060-02R	LIQUID	1	n/a		% Ct. Error
8 DUP	S98T003008	0	GEA-01 CS13702	LIQUID	2.15e+02	2.14e+02	0.466	RFD
8 DUP	S98T003008	0	GEA-01 CS13702R	LIQUID	1	0.0900	0.090	% Ct. Error
9 SAMPLE	S98T003009	0	GEA-01 C060-02	LIQUID	N/A	< 3.121e-03	3.12e-003	uCi/mL
9 SAMPLE	S98T003009	0	GEA-01 C060-02R	LIQUID	N/A	n/a		% Ct. Error
9 SAMPLE	S98T003009	0	GEA-01 CS13702	LIQUID	N/A	2.130e+02		uCi/mL
9 SAMPLE	S98T003009	0	GEA-01 CS13702R	LIQUID	N/A	0.0900		% Ct. Error
10 DUP	S98T003009	0	GEA-01 C060-02	LIQUID	<3.12e-3	<2.90e-3		RFD
10 DUP	S98T003009	0	GEA-01 C060-02R	LIQUID	1	n/a		% Ct. Error
10 DUP	S98T003009	0	GEA-01 CS13702	LIQUID	2.13e+02	2.11e+02	0.943	RFD
10 DUP	S98T003009	0	GEA-01 CS13702R	LIQUID	1	0.0900	0.090	% Ct. Error

Comments Section:

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26841

Seq	Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit

Comments for sample# S98T003006 and test @GEA-01 .

DL=0 => n/a.
DL=0 => n/a.

Comments for sample# S98T003007 and test @GEA-01 .

DL=0 => n/a.
DL=0 => n/a.

Comments for sample# S98T003008 and test @GEA-01 .

DL=0 => n/a.
DL=0 => n/a.

Comments for sample# S98T003009 and test @GEA-01 .

DL=0 => n/a.
DL=0 => n/a.

Final page for worklist# 26841

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

11-30-98
Reviewer Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26841

Analyst: gll Instrument: GEA00 #1 Book# 28657

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: xBY-105 FOR @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	LIQUID		
2	BLNK				@GEA-01	LIQUID		
3	SAMPLE	S98T003006 0			@GEA-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T003006 0			@GEA-01	LIQUID		
5	SAMPLE	S98T003007 0			@GEA-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6	DUP	S98T003007 0			@GEA-01	LIQUID		
7	SAMPLE	S98T003008 0			@GEA-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8	DUP	S98T003008 0			@GEA-01	LIQUID		
9	SAMPLE	S98T003009 0			@GEA-01	LIQUID	98000510	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
10	DUP	S98T003009 0			@GEA-01	LIQUID		

Final page for worklist # 26841

gll 11-23-98
Analyst Signature Date

[Signature] 11/25/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 14:30:38.67 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
 Sample ID: WL26841-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

844

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEAl
 File Number: dka300:[spec.GEAl]1g1125.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:14.50 sec
 Dead Time: 0.5%

Verified by:

RA Jones 11-25-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 13:39:48.45
 Decayed to: 23-NOV-1998 13:39:48.45
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEAl]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.85	77	0.65	25.82	24	6 98.5			
0	31.92	1120	0.88	63.96	59	10 12.3			
0	36.39	288	0.92	72.90	69	8 34.2			
0	59.71	224	0.88	119.53	116	8 47.2		AM-241	0.155
0	661.65*	6258	1.41	1323.40	1315	16 2.9		CS-137	7.58
0	1173.12	2766	1.84	2346.53	2337	20 4.5		CO-60	4.76
0	1332.30	2550	1.97	2664.99	2656	18 4.4		CO-60	4.86
0	1377.75*	13	1.56	2755.94	2752	9106.8			

Summary of Nuclide Activity
Sample ID : WL26841-STD

Page : 2
Acquisition date : 23-NOV-1998 13:39:48

Total number of lines in spectrum 8
Number of unidentified lines 4
Number of lines tentatively identified by NID 4 50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.813E+00	4.813E+00	0.151E+00	3.14	
CS-137	30.00Y	1.00	7.577E+00	7.577E+00	0.216E+00	2.86	
AM-241	432.20Y	1.00	1.550E-01	1.550E-01	0.732E-01	47.24	
Total Activity :			1.255E+01	1.255E+01			

Grand Total Activity : 1.255E+01 1.255E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26841-STD

Acquisition date : 23-NOV-1998 13:39:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	417.	477.59	7.1898E-01
NA-24	16.	1368.55	4.2759E-02
K-40	560.	1460.75	2.1641E+00
AR-41	46.	1293.60	7.5239E-02
SC-46	172.	1120.55	1.0535E-01
CR-51	451.	320.08	5.3614E-01
MN-54	192.	834.83	8.5875E-02
CO-56	183.	846.76	8.4921E-02
CO-58	190.	810.78	8.3568E-02
FE-59	173.	1099.25	1.8407E-01
ZN-65	190.	1115.55	2.1717E-01
SE-75	539.	264.66	8.6024E-02
KR-85	308.	514.00	1.5835E+01
SR-85	307.	514.01	7.1574E-02
Y-88	7.	1836.06	3.6148E-02
Y-91	70.	1204.67	2.4326E+01
NB-94	193.	871.09	8.9414E-02
NB-95	201.	765.78	8.1262E-02
ZR-95	182.	756.73	1.4023E-01
ZRNB-95	178.	724.18	3.3346E-01
RU-103	287.	497.08	7.0876E-02
RURH-106	205.	621.93	1.3620E+00
AG-108m	172.	722.94	7.8638E-02
CD-109	583.	88.03	8.0133E-01
AG-110M	316.	657.76	9.2544E-02
SN-113	431.	391.69	9.8669E-02
TE-123m	666.	159.00	4.6520E-02
SB-124	201.	602.73	6.6445E-02
SB-125	392.	427.89	2.2209E-01
TE-125m	647.	109.27	1.1450E+01
I-129	489.	39.60	4.0368E-01
I-131	379.	364.48	6.8538E-02
XE-131m	705.	163.93	2.0886E+00
BA-133	438.	356.02	9.4127E-02
CS-134	200.	604.70	6.6498E-02
CS-136	192.	818.51	8.4562E-02
CS-138	24.	1435.86	1.0957E-01
CE-139	686.	165.85	5.0858E-02
BA-140	278.	537.31	2.7982E-01
LA-140	19.	1596.21	5.3107E-02
CE-141	755.	145.44	8.2036E-02
CE-144	726.	133.51	3.3470E-01
CEPR-144	725.	133.51	6.6867E-01
EU-152	726.	121.78	1.2544E-01
EU-154	39.	1274.51	1.7011E-01
EU-155	627.	86.54	9.1122E-02
HF-181	342.	482.18	8.4748E-02
TA-182	524.	67.75	6.5219E-02
HG-203	493.	279.20	6.1777E-02
BI-207	200.	569.70	6.2817E-02
TL-208	519.	277.36	8.1364E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26841-STD

Acquisition date : 23-NOV-1998 13:39:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	466.	46.50	6.7734E-01
BI-212	185.	727.18	1.1173E+00
PB-212	674.	238.63	1.1873E-01
BI-214	368.	609.31	1.9595E-01
PB-214	503.	351.92	3.0872E-01
RA-224	616.	240.99	1.2616E+00
RA-226	785.	186.10	1.3335E+00
AC-228	312.	911.21	4.4095E-01
TH-228	609.	84.37	2.3164E+00
TH-229	599.	88.47	1.1876E-01
U-232	585.	57.78	1.4395E+01
PA-233	474.	312.17	1.4040E-01
UTH-233	555.	245.34	4.3707E+01
PA-234M	186.	1001.03	1.5269E+01
TH-234	532.	63.29	7.1463E-01
U-235	775.	185.71	8.0558E-02
NP-237	629.	86.48	2.4135E-01
U-237	801.	59.54	9.7057E-02
NP-238	189.	984.45	3.5652E-01
NP-239	692.	106.12	1.4498E-01
PU-239	716.	129.30	5.6673E+02
AM-243	648.	74.67	4.9733E-02

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 17:24:25.43 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
 Sample ID: WL26841-BLNK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]lg1126.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:33.23 sec
 Dead Time: 0.4%

Verified by:

11-25-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 14:53:20.02
 Decayed to: 23-NOV-1998 14:53:20.02
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.85	174	1.17	25.83	24	5 53.3			
0	16.00	201	0.94	32.12	29	7 60.2			
0	87.56*	131	1.22	175.23	173	7101.7			
								NP-237	8.90
								CD-109	30.7
								CS-137	8.51
0	661.53*	209	1.47	1323.16	1318	13 40.0			
0	1510.60	47	1.75	3021.72	3015	16 55.0			
0	1557.58	13	1.09	3115.72	3111	7 91.3			
0	1621.61	49	1.82	3243.84	3238	16 39.1			

Summary of Nuclide Activity

Page : 2

Sample ID : WL26841-BLNK

Acquisition date : 23-NOV-1998 14:53:20

Total number of lines in spectrum 7
 Number of unidentified lines 4
 Number of lines tentatively identified by NID 3 42.86%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CD-109~~	~~462.90D~~	~~1.00~~	~~3.068E+01~~	~~3.069E+01~~	~~3.121E+01~~	~~101.71~~	~~RAJ~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~8.514E+00~~	~~8.514E+00~~	~~3.406E+00~~	~~40.00~~	
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~8.904E+00~~	~~8.904E+00~~	~~9.056E+00~~	~~101.71~~	~~RAJ~~
Total Activity :			4.810E+01	4.811E+01			

11-25-98

Grand Total Activity : 4.810E+01 4.811E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26841-BLNK

Acquisition date : 23-NOV-1998 14:53:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	335.	477.59	2.1790E+01
NA-24	48.	1368.55	2.4351E+00
K-40	1735.	1460.75	1.2693E+02
AR-41	95.	1293.60	4.7081E+00
SC-46	261.	1120.55	4.3385E+00
CR-51	591.	320.08	2.0612E+01
MN-54	171.	834.83	2.7286E+00
CO-56	171.	846.76	2.7706E+00
CO-58	160.	810.78	2.5956E+00
FE-59	129.	1099.25	5.3928E+00
CO-60	72.	1332.50	2.7067E+00
ZN-65	126.	1115.55	5.9982E+00
SE-75	832.	264.66	3.5830E+00
KR-85	388.	514.00	5.9629E+02
SR-85	387.	514.01	2.6940E+00
Y-88	27.	1836.06	2.2072E+00
Y-91	142.	1204.67	1.1465E+03
NB-94	172.	871.09	2.8485E+00
NB-95	237.	765.78	2.9627E+00
ZR-95	191.	756.73	4.8324E+00
ZRNB-95	234.	724.18	1.2790E+01
RU-103	319.	497.08	2.5097E+00
RURH-106	228.	621.93	4.8191E+01
AG-108m	203.	722.94	2.8655E+00
AG-110M	225.	657.76	2.6442E+00
SN-113	426.	391.69	3.3026E+00
TE-123m	1345.	159.00	2.2112E+00
SB-124	273.	602.73	2.5902E+00
SB-125	420.	427.89	7.7286E+00
TE-125m	1397.	109.27	5.6235E+02
I-129	760.	39.60	1.6856E+01
I-131	508.	364.48	2.6714E+00
XE-131m	1325.	163.93	9.6015E+01
BA-133	497.	356.02	3.3721E+00
CS-134	252.	604.70	2.5025E+00
CS-136	173.	818.51	2.7155E+00
CS-138	54.	1435.86	1.1061E+01
CE-139	1304.	165.85	2.3479E+00
BA-140	283.	537.31	9.5167E+00
LA-140	32.	1596.21	2.3014E+00
CE-141	1329.	145.44	3.6475E+00
CE-144	1425.	133.51	1.5699E+01
CEPR-144	1426.	133.51	3.1372E+01
EU-152	1431.	121.78	5.8910E+00
EU-154	103.	1274.51	8.9955E+00
EU-155	1301.	86.54	4.3875E+00
HF-181	337.	482.18	2.8318E+00
TA-182	928.	67.75	2.9032E+00
HG-203	791.	279.20	2.6219E+00
BI-207	224.	569.70	2.2335E+00
TL-208	787.	277.36	3.3566E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26841-BLNK

Acquisition date : 23-NOV-1998 14:53:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	849.	46.50	3.0669E+01
BI-212	263.	727.18	4.4513E+01
PB-212	1284.	238.63	5.4841E+00
BI-214	682.	609.31	8.9124E+00
PB-214	860.	351.92	3.0062E+01
RA-224	1043.	240.99	5.5333E+01
RA-226	1247.	186.10	5.6365E+01
AC-228	367.	911.21	1.6062E+01
TH-228	1330.	84.37	1.1440E+02
TH-229	1264.	88.47	5.7673E+00
U-232	1052.	57.78	6.4616E+02
PA-233	606.	312.17	5.3366E+00
UTH-233	827.	245.34	1.7884E+03
PA-234M	167.	1001.03	4.8838E+02
TH-234	1144.	63.29	3.5018E+01
U-235	1248.	185.71	3.4267E+00
U-237	1071.	59.54	3.7800E+00
NP-238	140.	984.45	1.0539E+01
NP-239	1328.	106.12	6.7895E+00
PU-239	1506.	129.30	2.7490E+04
AM-241	1070.	59.54	3.6131E+00
AM-243	1338.	74.67	2.3898E+00

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 20:20:48.35 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26841
 Sample ID: S98T3006-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1127.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:53:58.54 sec
 Dead Time: 14%

Verified by:

11-25-98

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-NOV-1998 17:26:21.17
 Decayed to: 23-NOV-1998 17:26:21.17
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
2	8.94	21501	0.89	18.01	16	7	1.9	5.80E+02		
2	9.84	10741	0.89	19.81	16	7	6.6			
0	12.50	3309	0.82	25.12	22	7	34.5			
0	21.94	26828	1.26	44.01	41	7	4.8			
0	31.97	889805	1.23	64.06	59	11	0.4			
0	36.50	216791	1.21	73.11	69	8	0.9			
0	75.26*	1597	0.93	150.62	148	7	72.7		AM-243	22.1
0	84.65*	1045	1.23	169.41	167	6	97.8		TH-228	694.
0	187.95	21479	4.88	375.98	370	15	12.5			
2	651.77	3236	1.82	1303.63	1297	38	19.9	2.80E+03		
2	661.69*	5141079	1.71	1323.47	1297	38	0.1		CS-137	2.096E+05
0	692.62	2458	2.49	1385.34	1378	16	14.6			
0	1322.91	1059	1.96	2646.22	2638	19	8.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3006-SAM

Acquisition date : 23-NOV-1998 17:26:21

Total number of lines in spectrum	13	
Number of unidentified lines	10	
Number of lines tentatively identified by NID	3	23.08%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
CS-137	30.00Y	1.00	2.096E+05	2.096E+05	0.002E+05	0.09			
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~6.941E+02~~	~~6.941E+02~~	~~6.786E+02~~	~~97.76~~			
~~AM-243~~	~~7380.00Y*****~~		~~2.208E+01~~	~~2.208E+01~~	~~1.606E+01~~	~~72.71~~			
Total Activity :			2.103E+05	2.103E+05					

11-25-98

Grand Total Activity :	2.103E+05	2.103E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3006-SAM

Acquisition date : 23-NOV-1998 17:26:21

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	164497.	477.59	4.6849E+02
NA-24	52.	1368.55	2.5403E+00
K-40	1745.	1460.75	1.2726E+02
AR-41	87.	1293.60	4.8092E+00
SC-46	409.	1120.55	5.3946E+00
CR-51	125935.	320.08	2.9444E+02
MN-54	322.	834.83	3.7081E+00
CO-56	331.	846.76	3.8096E+00
CO-58	303.	810.78	3.5255E+00
FE-59	255.	1099.25	7.4727E+00
CO-60	83.	1332.50	2.8978E+00
ZN-65	263.	1115.55	8.5498E+00
SE-75	126913.	264.66	4.3444E+01
KR-85	80978.	514.00	8.3891E+03
SR-85	80969.	514.01	3.7948E+01
Y-88	20.	1836.06	1.9069E+00
Y-91	179.	1204.67	1.2819E+03
NB-94	298.	871.09	3.7143E+00
NB-95	351.	765.78	3.5823E+00
ZR-95	358.	756.73	6.5481E+00
ZRNB-95	435.	724.18	1.7274E+01
RU-103	98056.	497.08	4.2719E+01
RURH-106	24876.	621.93	4.8667E+02
AG-108m	421.	722.94	4.0747E+00
CD-109	99153.	88.03	3.4418E+02
AG-110M	194931.	657.76	7.5077E+01
SN-113	144163.	391.69	5.9174E+01
TE-123m	115476.	159.00	2.0202E+01
SB-124	26758.	602.73	2.4874E+01
SB-125	168601.	427.89	1.5084E+02
TE-125m	105947.	109.27	4.8302E+03
I-129	92882.	39.60	1.8285E+02
I-131	133498.	364.48	4.2296E+01
XE-131m	116485.	163.93	8.8817E+02
BA-133	132303.	356.02	5.3691E+01
CS-134	26525.	604.70	2.4876E+01
CS-136	339.	818.51	3.7537E+00
CS-138	55.	1435.86	1.2680E+01
CE-139	116453.	165.85	2.1871E+01
BA-140	57062.	537.31	1.3106E+02
LA-140	27.	1596.21	2.1250E+00
CE-141	113922.	145.44	3.3308E+01
CE-144	114870.	133.51	1.3904E+02
CEPR-144	114871.	133.51	2.7783E+02
EU-152	113992.	121.78	5.1877E+01
EU-154	104.	1274.51	9.0127E+00
EU-155	99404.	86.54	3.7807E+01
HF-181	128733.	482.18	5.3779E+01
TA-182	93555.	67.75	2.8664E+01
HG-203	126703.	279.20	3.2571E+01
BI-207	33201.	569.70	2.6262E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3006-SAM

Acquisition date : 23-NOV-1998 17:26:21

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	126573.	277.36	4.1765E+02
PB-210	90054.	46.50	3.1044E+02
BI-212	504.	727.18	6.1074E+01
PB-212	131536.	238.63	5.4701E+01
BI-214	26226.	609.31	5.4267E+01
PB-214	131882.	351.92	4.1849E+02
RA-224	130869.	240.99	6.1073E+02
RA-226	174843.	186.10	6.5748E+02
AC-228	512.	911.21	1.8879E+01
TH-229	98963.	88.47	5.0299E+01
U-232	90897.	57.78	5.9115E+03
PA-233	126021.	312.17	7.5293E+01
UTH-233	129925.	245.34	2.2012E+04
PA-234M	247.	1001.03	5.8974E+02
TH-234	93122.	63.29	3.1117E+02
U-235	172230.	185.71	3.9667E+01
NP-237	99485.	86.48	1.0003E+02
U-237	91669.	59.54	3.4463E+01
NP-238	234.	984.45	1.3482E+01
NP-239	103049.	106.12	5.9119E+01
PU-239	114322.	129.30	2.3635E+05
AM-241	91670.	59.54	3.2914E+01

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 23:16:05.91
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26841
 Sample ID: S98T003006-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEAL
 File Number: dka300:[spec.GEAL]1g1128.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:53:50.07 sec
 Dead Time: 14%

Verified by:

R. Jones 11-25-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-NOV-1998 20:21:47.92
 Decayed to: 23-NOV-1998 20:21:47.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEAL]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	9.26	34271	1.44	18.65	16	7	2.5			
0	12.54	2707	0.76	25.21	22	8	46.8			
0	21.97	25769	1.26	44.06	41	7	5.0			
0	31.98	885870	1.23	64.09	59	11	0.4			
0	36.52	215100	1.20	73.15	69	8	0.9			
0	75.13*	2068	1.06	150.36	148	7	55.9		AM-243	28.6
0	84.77*	1548	0.95	169.66	167	7	74.8		TH-228	1.028E+03
0	187.94	24480	4.79	375.96	370	15	10.9			
0	661.66*	5158473	1.64	1323.41	1315	17	0.1		CS-137	2.103E+05
0	692.69	2057	2.29	1385.49	1379	13	14.8			
0	697.45	273	1.01	1395.00	1392	8	47.8			
0	1322.88	1125	2.03	2646.15	2639	18	8.5			

Summary of Nuclide Activity

Sample ID : S98T003006-DUP

Page : 2

Acquisition date : 23-NOV-1998 20:21:47

Total number of lines in spectrum 12
 Number of unidentified lines 9
 Number of lines tentatively identified by NID 3 25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.103E+05	2.103E+05	0.002E+05	0.09	
TH-228	1.91Y	1.00	1.028E+03	1.028E+03	0.769E+03	74.80	RMF
AM-243	7380.00Y*****		2.860E+01	2.860E+01	1.600E+01	55.94	RMF
Total Activity :			2.113E+05	2.113E+05			

11-25-78

Grand Total Activity : 2.113E+05 2.113E+05

Flags: "K" = Keyline not found

"E" = Manually edited

"M" = Manually accepted

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T003006-DUP

Acquisition date : 23-NOV-1998 20:21:47

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	163720.	477.59	4.6738E+02
NA-24	54.	1368.55	2.5824E+00
K-40	1683.	1460.75	1.2501E+02
AR-41	95.	1293.60	4.9897E+00
SC-46	396.	1120.55	5.3098E+00
CR-51	126278.	320.08	2.9484E+02
MN-54	290.	834.83	3.5235E+00
CO-56	339.	846.76	3.8546E+00
CO-58	301.	810.78	3.5114E+00
FE-59	242.	1099.25	7.2863E+00
CO-60	93.	1332.50	3.0502E+00
ZN-65	237.	1115.55	8.1369E+00
SE-75	126782.	264.66	4.3421E+01
KR-85	80395.	514.00	8.3589E+03
SR-85	80387.	514.01	3.7812E+01
Y-88	20.	1836.06	1.9165E+00
Y-91	192.	1204.67	1.3263E+03
NB-94	301.	871.09	3.7310E+00
NB-95	305.	765.78	3.3453E+00
ZR-95	359.	756.73	6.5585E+00
ZRNB-95	426.	724.18	1.7099E+01
RU-103	98677.	497.08	4.2854E+01
RURH-106	24990.	621.93	4.8779E+02
AG-108m	407.	722.94	4.0093E+00
CD-109	98225.	88.03	3.4257E+02
AG-110M	196734.	657.76	7.5423E+01
SN-113	142661.	391.69	5.8865E+01
TE-123m	114950.	159.00	2.0156E+01
SB-124	26327.	602.73	2.4673E+01
SB-125	168062.	427.89	1.5060E+02
TE-125m	106048.	109.27	4.8325E+03
I-129	92173.	39.60	1.8215E+02
I-131	133435.	364.48	4.2286E+01
XE-131m	116256.	163.93	8.8729E+02
BA-133	130958.	356.02	5.3418E+01
CS-134	26022.	604.70	2.4639E+01
CS-136	311.	818.51	3.6020E+00
CS-138	57.	1435.86	1.2960E+01
CE-139	116912.	165.85	2.1914E+01
BA-140	56957.	537.31	1.3094E+02
LA-140	53.	1596.21	2.8929E+00
CE-141	114363.	145.44	3.3373E+01
CE-144	113560.	133.51	1.3825E+02
CEPR-144	113560.	133.51	2.7625E+02
EU-152	112860.	121.78	5.1620E+01
EU-154	131.	1274.51	1.0077E+01
EU-155	98340.	86.54	3.7605E+01
HF-181	127726.	482.18	5.3568E+01
TA-182	93035.	67.75	2.8584E+01
HG-203	124974.	279.20	3.2348E+01
BI-207	32927.	569.70	2.6153E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T003006-DUP

Acquisition date : 23-NOV-1998 20:21:47

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	125757.	277.36	4.1631E+02
PB-210	90429.	46.50	3.1109E+02
BI-212	459.	727.18	5.8328E+01
PB-212	130697.	238.63	5.4527E+01
BI-214	26214.	609.31	5.4255E+01
PB-214	130771.	351.92	4.1640E+02
RA-224	129651.	240.99	6.0788E+02
RA-226	174179.	186.10	6.5623E+02
AC-228	543.	911.21	1.9425E+01
TH-229	98331.	88.47	5.0138E+01
U-232	91152.	57.78	5.9197E+03
PA-233	125204.	312.17	7.5049E+01
UTH-233	129005.	245.34	2.1934E+04
PA-234M	253.	1001.03	5.9674E+02
TH-234	93129.	63.29	3.1118E+02
U-235	171352.	185.71	3.9566E+01
NP-237	98469.	86.48	9.9519E+01
U-237	90957.	59.54	3.4329E+01
NP-238	243.	984.45	1.3741E+01
NP-239	103007.	106.12	5.9107E+01
PU-239	113618.	129.30	2.3562E+05
AM-241	90957.	59.54	3.2786E+01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 02:16:02.56 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26841
 Sample ID: S98T003007-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

Raymond

>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1129.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:52:25.27 sec
 Dead Time: 13%

Verified by:

Ed Jones 11-25-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-NOV-1998 23:23:07.77
 Decayed to: 23-NOV-1998 23:23:07.77
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	9.24	31795	1.42	18.61	16	7	2.7			
4	11.71	745	0.71	23.53	22	6	58.2	2.56E+00		
4	12.57	2031	0.57	25.27	22	6	28.7			
0	21.97	25321	1.28	44.06	41	7	4.9			
0	31.97	825126	1.21	64.06	59	10	0.4			
0	36.54	222812	1.22	73.19	69	9	0.8			
0	75.21*	1845	0.94	150.52	148	7	60.9		AM-243	25.5
0	84.55*	1207	0.97	169.21	168	7	92.8		TH-228	802.
0	187.72	19069	3.84	375.53	370	14	12.8			
0	651.05	630	1.16	1302.21	1301	7	81.7			
0	661.66*	4864278	1.62	1323.42	1315	17	0.1		CS-137	1.983E+05
0	692.66	2017	2.51	1385.42	1379	15	15.6			
0	1322.89	965	1.95	2646.16	2639	18	9.0			
0	1511.06	31	0.78	3022.66	3014	16	101.4			
0	1730.40	39	1.35	3461.53	3456	13	41.6			

Summary of Nuclide Activity
Sample ID : S98T003007-SAM

Page : 2
Acquisition date : 23-NOV-1998 23:23:07

Total number of lines in spectrum 15
Number of unidentified lines 12
Number of lines tentatively identified by NID 3 20.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error Flags
CS-137	30.00Y	1.00	1.983E+05	1.983E+05	0.002E+05	0.09
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~8.018E+02~~	~~8.019E+02~~	~~7.445E+02~~	~~92.85~~
~~AM-243~~	~~7380.00Y*****~~	~~-----~~	~~2.551E+01~~	~~2.551E+01~~	~~1.553E+01~~	~~60.90~~
Total Activity :			1.991E+05	1.991E+05		

11-25-98

Grand Total Activity : 1.991E+05 1.991E+05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
 Sample ID : S98T003007-SAM

Page : 3
 Acquisition date : 23-NOV-1998 23:23:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	154099.	477.59	4.5346E+02
NA-24	52.	1368.55	2.5416E+00
K-40	1748.	1460.75	1.2740E+02
AR-41	121.	1293.60	5.5874E+00
SC-46	365.	1120.55	5.1022E+00
CR-51	118376.	320.08	2.8548E+02
MN-54	294.	834.83	3.5443E+00
CO-56	311.	846.76	3.6950E+00
CO-58	313.	810.78	3.5828E+00
FE-59	261.	1099.25	7.5579E+00
CO-60	111.	1332.50	3.3228E+00
ZN-65	236.	1115.55	8.1085E+00
SE-75	119905.	264.66	4.2229E+01
KR-85	75874.	514.00	8.1210E+03
SR-85	75865.	514.01	3.6735E+01
Y-88	19.	1836.06	1.8902E+00
Y-91	166.	1204.67	1.2376E+03
NB-94	278.	871.09	3.5918E+00
NB-95	348.	765.78	3.5668E+00
ZR-95	341.	756.73	6.3920E+00
ZRNB-95	440.	724.18	1.7377E+01
RU-103	92886.	497.08	4.1580E+01
RURH-106	23520.	621.93	4.7328E+02
AG-108m	425.	722.94	4.0937E+00
CD-109	92579.	88.03	3.3260E+02
AG-110M	180636.	657.76	7.2275E+01
SN-113	135075.	391.69	5.7281E+01
TE-123m	108584.	159.00	1.9591E+01
SB-124	24896.	602.73	2.3995E+01
SB-125	158905.	427.89	1.4645E+02
TE-125m	99738.	109.27	4.6868E+03
I-129	86055.	39.60	1.7602E+02
I-131	125538.	364.48	4.1016E+01
XE-131m	108521.	163.93	8.5729E+02
BA-133	123694.	356.02	5.1918E+01
CS-134	24654.	604.70	2.3985E+01
CS-136	301.	818.51	3.5442E+00
CS-138	39.	1435.86	1.0820E+01
CE-139	109162.	165.85	2.1176E+01
BA-140	54141.	537.31	1.2767E+02
LA-140	50.	1596.21	2.8136E+00
CE-141	107433.	145.44	3.2347E+01
CE-144	107571.	133.51	1.3456E+02
CEPR-144	107570.	133.51	2.6887E+02
EU-152	107401.	121.78	5.0358E+01
EU-154	129.	1274.51	1.0010E+01
EU-155	93027.	86.54	3.6576E+01
HF-181	121172.	482.18	5.2178E+01
TA-182	87659.	67.75	2.7747E+01
HG-203	118223.	279.20	3.1464E+01
BI-207	31203.	569.70	2.5462E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T003007-SAM

Acquisition date : 23-NOV-1998 23:23:07

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	118644.	277.36	4.0438E+02
PB-210	85133.	46.50	3.0185E+02
BI-212	489.	727.18	6.0181E+01
PB-212	123590.	238.63	5.3026E+01
BI-214	24396.	609.31	5.2347E+01
PB-214	123314.	351.92	4.0126E+02
RA-224	123284.	240.99	5.9273E+02
RA-226	164046.	186.10	6.3689E+02
AC-228	490.	911.21	1.8484E+01
TH-229	92742.	88.47	4.8695E+01
U-232	85780.	57.78	5.7430E+03
PA-233	117739.	312.17	7.2780E+01
UTH-233	121309.	245.34	2.1271E+04
PA-234M	228.	1001.03	5.6660E+02
TH-234	86971.	63.29	3.0074E+02
U-235	161555.	185.71	3.8419E+01
NP-237	93115.	86.48	9.6781E+01
U-237	86439.	59.54	3.3465E+01
NP-238	224.	984.45	1.3210E+01
NP-239	96756.	106.12	5.7280E+01
PU-239	108053.	129.30	2.2979E+05
AM-241	86439.	59.54	3.1963E+01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 09:26:38.96 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
 Sample ID: S98T3007-DUP
 Sample Size: 1.0000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

Sh L Usher

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1132.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:52:36.93 sec
 Dead Time: 13%

Verified by:

RA Jones 11-25-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 24-NOV-1998 06:33:30.78
 Decayed to: 24-NOV-1998 06:33:30.78
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	9.25	32437	1.49	18.62	16	7	2.6			
0	12.45	2632	0.68	25.03	22	8	47.1			
0	21.93	25099	1.32	43.99	41	7	5.0			
0	31.97	852818	1.21	64.05	59	11	0.4			
0	36.53	227151	1.24	73.18	69	9	0.8			
0	75.01*	1950	1.03	150.12	148	7	57.8		AM-243	27.0
0	84.97*	958	1.39	170.04	168	6(104.1)			TH-228	636.
0	187.80	20120	3.82	375.69	370	15	13.0			
0	661.69*	4918004	1.62	1323.47	1315	17	0.1		CS-137	2.005E+05
0	692.44	2019	3.19	1384.98	1379	15	15.8			
0	697.62	220	2.43	1395.35	1393	10	63.8			
0	1322.92	1038	1.97	2646.24	2638	17	7.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3007-DUP

Acquisition date : 24-NOV-1998 06:33:30

Total number of lines in spectrum 12
 Number of unidentified lines 9
 Number of lines tentatively identified by NID 3 25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error Flags
CS-137	30.00Y	1.00	2.005E+05	2.005E+05	0.002E+05	0.09
TH-228	1.91Y	1.00	6.360E+02	6.360E+02	6.623E+02	104.13
AM-243	7380.00Y*****		2.696E+01	2.696E+01	1.557E+01	57.76
Total Activity :			2.011E+05	2.011E+05		

11-25-98

Grand Total Activity : 2.011E+05 2.011E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3007-DUP

Acquisition date : 24-NOV-1998 06:33:30

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	156009.	477.59	4.5625E+02
NA-24	56.	1368.55	2.6200E+00
K-40	1771.	1460.75	1.2819E+02
AR-41	116.	1293.60	5.4889E+00
SC-46	367.	1120.55	5.1166E+00
CR-51	120010.	320.08	2.8744E+02
MN-54	292.	834.83	3.5349E+00
CO-56	364.	846.76	3.9872E+00
CO-58	287.	810.78	3.4306E+00
FE-59	245.	1099.25	7.3343E+00
CO-60	81.	1332.50	2.8580E+00
ZN-65	249.	1115.55	8.3261E+00
SE-75	120402.	264.66	4.2317E+01
KR-85	76354.	514.00	8.1466E+03
SR-85	76344.	514.01	3.6851E+01
Y-88	22.	1836.06	2.0027E+00
Y-91	184.	1204.67	1.2988E+03
NB-94	302.	871.09	3.7367E+00
NB-95	315.	765.78	3.3975E+00
ZR-95	356.	756.73	6.5304E+00
ZRNB-95	373.	724.18	1.6028E+01
RU-103	93279.	497.08	4.1667E+01
RURH-106	23600.	621.93	4.7407E+02
AG-108m	399.	722.94	3.9702E+00
CD-109	93974.	88.03	3.3509E+02
AG-110M	176352.	657.76	7.1414E+01
SN-113	136118.	391.69	5.7501E+01
TE-123m	109828.	159.00	1.9703E+01
SB-124	24773.	602.73	2.3936E+01
SB-125	160389.	427.89	1.4713E+02
TE-125m	100252.	109.27	4.6988E+03
I-129	87503.	39.60	1.7749E+02
I-131	127043.	364.48	4.1261E+01
XE-131m	109489.	163.93	8.6110E+02
BA-133	123961.	356.02	5.1974E+01
CS-134	24667.	604.70	2.3992E+01
CS-136	283.	818.51	3.4374E+00
CS-138	50.	1435.86	1.2062E+01
CE-139	109811.	165.85	2.1239E+01
BA-140	54146.	537.31	1.2767E+02
LA-140	28.	1596.21	2.1652E+00
CE-141	108022.	145.44	3.2435E+01
CE-144	107815.	133.51	1.3471E+02
CEPR-144	107815.	133.51	2.6918E+02
EU-152	108005.	121.78	5.0499E+01
EU-154	109.	1274.51	9.2427E+00
EU-155	94088.	86.54	3.6784E+01
HF-181	122707.	482.18	5.2507E+01
TA-182	88321.	67.75	2.7852E+01
HG-203	118999.	279.20	3.1567E+01
BI-207	31229.	569.70	2.5472E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3007-DUP

Acquisition date : 24-NOV-1998 06:33:30

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	119634.	277.36	4.0606E+02
PB-210	85773.	46.50	3.0298E+02
BI-212	442.	727.18	5.7262E+01
PB-212	123561.	238.63	5.3020E+01
BI-214	24485.	609.31	5.2443E+01
PB-214	124767.	351.92	4.0404E+02
RA-224	123335.	240.99	5.9286E+02
RA-226	165510.	186.10	6.3972E+02
AC-228	469.	911.21	1.8086E+01
TH-229	94039.	88.47	4.9034E+01
U-232	86670.	57.78	5.7727E+03
PA-233	118666.	312.17	7.3065E+01
UTH-233	122979.	245.34	2.1417E+04
PA-234M	262.	1001.03	6.0653E+02
TH-234	88057.	63.29	3.0260E+02
U-235	163004.	185.71	3.8591E+01
NP-237	94112.	86.48	9.7296E+01
U-237	87121.	59.54	3.3597E+01
NP-238	231.	984.45	1.3423E+01
NP-239	97063.	106.12	5.7372E+01
PU-239	107650.	129.30	2.2936E+05
AM-241	87121.	59.54	3.2089E+01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 12:51:37.95
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
 Sample ID: S98T3008-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]lg1133.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:54:35.48 sec
 Dead Time: 14%

Verified by:

11-25-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 24-NOV-1998 09:26:58.39
 Decayed to: 24-NOV-1998 09:26:58.39
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	9.26	35242	1.47	18.64	16	7	2.5			
0	12.61	3006	0.88	25.33	22	8	42.9			
0	21.99	28082	1.33	44.09	41	7	4.6			
0	31.98	923723	1.23	64.09	59	11	0.4			
0	36.51	222374	1.21	73.14	69	8	0.9			
0	75.03*	1952	1.00	150.18	148	7	60.0		AM-243	27.0
0	84.73*	2314	1.26	169.57	167	7	50.5		TH-228	1.536E+03
0	87.75*	1251	0.98	175.61	174	6	82.3		NP-237	85.2
									CD-109	294.
0	187.94	24507	4.49	375.96	370	16	11.7			
0	473.66	4014	2.15	947.40	944	8	37.8			
2	651.77	3780	1.82	1303.63	1296	40	17.1	2.74E+03		
2	661.72*	5277529	1.73	1323.54	1296	40	0.1		CS-137	2.151E+05
0	692.64	2820	3.15	1385.40	1378	16	13.0			
0	1322.84	1239	2.09	2646.07	2639	17	7.3			
0	1661.23	18	2.26	3323.12	3318	11	85.0			

Summary of Nuclide Activity
Sample ID : S98T3008-SAM

Page : 2
Acquisition date : 24-NOV-1998 09:26:58

Total number of lines in spectrum 15
Number of unidentified lines 11
Number of lines tentatively identified by NID 4 26.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~GB-109~~	~~462.90D~~	~~1.00~~	~~2.935E+02~~	~~2.935E+02~~	~~2.415E+02~~	~~82.25~~	~~RAJ~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~2.151E+05~~	~~2.151E+05~~	~~0.002E+05~~	~~0.09~~	~~RAJ~~
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~1.536E+03~~	~~1.536E+03~~	~~0.776E+03~~	~~50.48~~	~~RAJ~~
~~NP-237~~	~~2.14E+06Y~~	~~1.00~~	~~8.517E+01~~	~~8.517E+01~~	~~7.006E+01~~	~~82.25~~	~~RAJ~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~2.700E+01~~	~~2.700E+01~~	~~1.621E+01~~	~~60.03~~	~~RAJ~~
Total Activity :			2.171E+05	2.171E+05			11-25-98

Grand Total Activity : 2.171E+05 2.171E+05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3008-SAM

Acquisition date : 24-NOV-1998 09:26:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	169549.	477.59	4.7562E+02
NA-24	48.	1368.55	2.4512E+00
K-40	1748.	1460.75	1.2738E+02
AR-41	104.	1293.60	5.2383E+00
SC-46	416.	1120.55	5.4352E+00
CR-51	129358.	320.08	2.9841E+02
MN-54	334.	834.83	3.7718E+00
CO-56	341.	846.76	3.8651E+00
CO-58	326.	810.78	3.6535E+00
FE-59	242.	1099.25	7.2800E+00
CO-60	93.	1332.50	3.0472E+00
ZN-65	277.	1115.55	8.7698E+00
SE-75	129677.	264.66	4.3914E+01
KR-85	82822.	514.00	8.4839E+03
SR-85	82812.	514.01	3.8377E+01
Y-88	14.	1836.06	1.6377E+00
Y-91	153.	1204.67	1.1874E+03
NB-94	335.	871.09	3.9273E+00
NB-95	377.	765.78	3.7106E+00
ZR-95	380.	756.73	6.7366E+00
ZRNB-95	446.	724.18	1.7492E+01
RU-103	101353.	497.08	4.3430E+01
RURH-106	25390.	621.93	4.9166E+02
AG-108m	437.	722.94	4.1487E+00
AG-110M	193438.	657.76	7.4789E+01
SN-113	148530.	391.69	6.0062E+01
TE-123m	117943.	159.00	2.0416E+01
SB-124	26986.	602.73	2.4979E+01
SB-125	173489.	427.89	1.5301E+02
TE-125m	108742.	109.27	4.8934E+03
I-129	95067.	39.60	1.8499E+02
I-131	137620.	364.48	4.2944E+01
XE-131m	119462.	163.93	8.9943E+02
BA-133	134899.	356.02	5.4215E+01
CS-134	26606.	604.70	2.4913E+01
CS-136	314.	818.51	3.6167E+00
CS-138	52.	1435.86	1.2479E+01
CE-139	120019.	165.85	2.2203E+01
BA-140	58531.	537.31	1.3274E+02
LA-140	32.	1596.21	2.3106E+00
CE-141	116962.	145.44	3.3749E+01
CE-144	116603.	133.51	1.4009E+02
CEPR-144	116603.	133.51	2.7992E+02
EU-152	116362.	121.78	5.2413E+01
EU-154	108.	1274.51	9.1814E+00
EU-155	101105.	86.54	3.8129E+01
HF-181	132354.	482.18	5.4529E+01
TA-182	95586.	67.75	2.8973E+01
HG-203	128961.	279.20	3.2860E+01
BI-207	33988.	569.70	2.6570E+01
TL-208	129033.	277.36	4.2169E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3008-SAM

Acquisition date : 24-NOV-1998 09:26:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	92698.	46.50	3.1497E+02
BI-212	517.	727.18	6.1788E+01
PB-212	134300.	238.63	5.5272E+01
BI-214	26671.	609.31	5.4725E+01
PB-214	134230.	351.92	4.2361E+02
RA-224	133588.	240.99	6.1706E+02
RA-226	178618.	186.10	6.6453E+02
AC-228	514.	911.21	1.8912E+01
TH-229	101533.	88.47	5.0947E+01
U-232	93943.	57.78	6.0095E+03
PA-233	129170.	312.17	7.6227E+01
UTH-233	132901.	245.34	2.2263E+04
PA-234M	250.	1001.03	5.9342E+02
TH-234	95206.	63.29	3.1462E+02
U-235	175877.	185.71	4.0084E+01
U-237	94462.	59.54	3.4984E+01
NP-238	250.	984.45	1.3921E+01
NP-239	105225.	106.12	5.9743E+01
PU-239	116900.	129.30	2.3900E+05
AM-241	94463.	59.54	3.3411E+01

*****

* 222-S Laboratory Counting Room 24-NOV-1998 15:49:09.77 *

*****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
Sample ID: S98T3008-DUP
Sample Size: 1.00000E-03 L
Dilution Factor: 1.01000E+02

Removed by:

Shahid Uddin

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
File Number: dka300:[spec.GEA1]lg1135.cnf
Geometry: 41
Count Time: 0 02:30:00.00 sec
Real Time: 0 02:54:12.02 sec
Dead Time: 14%

Verified by:

R. Jones 11-25-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 24-NOV-1998 12:54:28.88
Decayed to: 24-NOV-1998 12:54:28.88
Standard Deviations: 2
Analysis Library: ENVGEA
Analyst: SZC
Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	9.25	34741	1.50	18.62	16	7	2.5			
3	11.64	666	0.65	23.40	22	8	67.4	1.28E+00		
3	12.61	2933	1.00	25.35	22	8	34.5			
0	21.95	26115	1.29	44.03	41	8	5.5			
0	31.98	907195	1.23	64.08	59	11	0.4			
0	36.51	218890	1.22	73.13	69	8	0.9			
0	75.11*	2378	1.14	150.32	148	6	43.3		AM-243	32.9
0	84.68*	1636	0.95	169.47	167	6	62.7		TH-228	1.087E+03
0	188.02	26115	4.35	376.13	370	16	10.9			
0	661.71*	5239173	1.63	1323.51	1315	18	0.1		CS-137	2.136E+05
0	692.55	2175	2.31	1385.20	1378	16	16.8			
0	1121.63	94	1.57	2243.52	2237	14	89.6			
0	1322.94	1191	1.97	2646.27	2636	19	7.5			
0	1775.09	17	0.60	3550.94	3543	17	107.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3008-DUP

Acquisition date : 24-NOV-1998 12:54:28

Total number of lines in spectrum 14
 Number of unidentified lines 10
 Number of lines tentatively identified by NID 4 28.57%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.136E+05	2.136E+05	0.002E+05	0.09	
TH-228	1.91Y	1.00	1.086E+03	1.087E+03	0.682E+03	62.74	
AM-243	7380.00Y	1.00	3.288E+01	3.288E+01	1.424E+01	43.30	
Total Activity :			2.147E+05	2.147E+05			

11-25-98

Grand Total Activity : 2.147E+05 2.147E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3008-DUP

Acquisition date : 24-NOV-1998 12:54:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	166574.	477.59	4.7143E+02
NA-24	65.	1368.55	2.8162E+00
K-40	1713.	1460.75	1.2611E+02
AR-41	133.	1293.60	5.8682E+00
SC-46	383.	1120.55	5.2218E+00
CR-51	127549.	320.08	2.9632E+02
MN-54	314.	834.83	3.6636E+00
CO-56	343.	846.76	3.8740E+00
CO-58	314.	810.78	3.5836E+00
FE-59	233.	1099.25	7.1563E+00
CO-60	99.	1332.50	3.1406E+00
ZN-65	275.	1115.55	8.7438E+00
SE-75	128554.	264.66	4.3724E+01
KR-85	81634.	514.00	8.4230E+03
SR-85	81625.	514.01	3.8102E+01
Y-88	30.	1836.06	2.3028E+00
Y-91	181.	1204.67	1.2878E+03
NB-94	317.	871.09	3.8224E+00
NB-95	364.	765.78	3.6479E+00
ZR-95	315.	756.73	6.1522E+00
ZRNB-95	449.	724.18	1.7550E+01
RU-103	100184.	497.08	4.3180E+01
RURH-106	25016.	621.93	4.8804E+02
AG-108m	450.	722.94	4.2073E+00
CD-109	99877.	88.03	3.4543E+02
AG-110M	188370.	657.76	7.3804E+01
SN-113	145697.	391.69	5.9487E+01
TE-123m	116229.	159.00	2.0268E+01
SB-124	26695.	602.73	2.4844E+01
SB-125	170321.	427.89	1.5161E+02
TE-125m	107535.	109.27	4.8662E+03
I-129	93244.	39.60	1.8321E+02
I-131	135158.	364.48	4.2559E+01
XE-131m	117315.	163.93	8.9132E+02
BA-133	132945.	356.02	5.3821E+01
CS-134	26402.	604.70	2.4818E+01
CS-136	300.	818.51	3.5359E+00
CS-138	47.	1435.86	1.1851E+01
CE-139	117994.	165.85	2.2015E+01
BA-140	57560.	537.31	1.3163E+02
LA-140	41.	1596.21	2.5847E+00
CE-141	115221.	145.44	3.3497E+01
CE-144	114588.	133.51	1.3887E+02
CEPR-144	114589.	133.51	2.7749E+02
EU-152	114534.	121.78	5.2000E+01
EU-154	124.	1274.51	9.8100E+00
EU-155	100161.	86.54	3.7951E+01
HF-181	131587.	482.18	5.4371E+01
TA-182	95160.	67.75	2.8908E+01
HG-203	126636.	279.20	3.2562E+01
BI-207	33283.	569.70	2.6294E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3008-DUP

Acquisition date : 24-NOV-1998 12:54:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	126368.	277.36	4.1731E+02
PB-210	91141.	46.50	3.1231E+02
BI-212	472.	727.18	5.9112E+01
PB-212	132358.	238.63	5.4872E+01
BI-214	26018.	609.31	5.4053E+01
PB-214	132211.	351.92	4.1952E+02
RA-224	132358.	240.99	6.1420E+02
RA-226	176364.	186.10	6.6033E+02
AC-228	480.	911.21	1.8302E+01
TH-229	99783.	88.47	5.0507E+01
U-232	92223.	57.78	5.9544E+03
PA-233	126202.	312.17	7.5347E+01
UTH-233	130923.	245.34	2.2097E+04
PA-234M	252.	1001.03	5.9475E+02
TH-234	93369.	63.29	3.1158E+02
U-235	173292.	185.71	3.9789E+01
NP-237	100260.	86.48	1.0042E+02
U-237	92460.	59.54	3.4611E+01
NP-238	251.	984.45	1.3963E+01
NP-239	103633.	106.12	5.9288E+01
PU-239	114648.	129.30	2.3669E+05
AM-241	92461.	59.54	3.3056E+01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 19:32:14.85 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26841
 Sample ID: S98T3009-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

Em Buca

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1136.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:54:18.20 sec
 Dead Time: 14%

Verified by:

WJmer 11-25-98

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 24-NOV-1998 16:37:26.92
 Decayed to: 24-NOV-1998 16:37:26.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	9.23	35041	1.51	18.59	16	7	2.5			
2	11.36	436	0.82	22.85	22	7	66.4	1.40E+01		
2	12.59	3143	0.82	25.31	22	7	28.0			
0	21.96	27767	1.30	44.05	41	7	4.7			
0	31.97	912288	1.23	64.06	59	11	0.4			
0	36.49	221520	1.21	73.11	69	8	0.9			
0	74.97*	1890	1.26	150.05	148	7	61.8		AM-243	26.1
0	84.81*	1208	1.56	169.72	167	7	96.6		TH-228	802.
0	188.00	24384	4.34	376.08	370	16	11.7			
2	651.77	3500	1.82	1303.63	1297	43	18.0	2.53E+03		
2	661.72*	5219547	1.72	1323.54	1297	43	0.1		CS-137	2.128E+05
0	693.15	1419	1.95	1386.41	1383	11	18.8			
0	697.79	292	1.54	1395.68	1393	10	51.3			
0	1323.06	1141	2.14	2646.50	2639	20	8.6			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3009-SAM

Acquisition date : 24-NOV-1998 16:37:26

Total number of lines in spectrum 14
 Number of unidentified lines 11
 Number of lines tentatively identified by NID 3 21.43%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error Flags
CS-137	30.00Y	1.00	2.128E+05	2.128E+05	0.002E+05	0.09
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~8.019E+02~~	~~8.019E+02~~	~~7.747E+02~~	~~96.61~~
~~AM-243~~	~~7380.00Y*****~~	~~1.00~~	~~2.613E+01~~	~~2.613E+01~~	~~1.614E+01~~	~~61.79~~
Total Activity :			2.136E+05	2.136E+05		

11-25-98

Grand Total Activity : 2.136E+05 2.136E+05

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3009-SAM

Acquisition date : 24-NOV-1998 16:37:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	167270.	477.59	4.7241E+02
NA-24	63.	1368.55	2.7747E+00
K-40	1721.	1460.75	1.2641E+02
AR-41	99.	1293.60	5.0944E+00
SC-46	409.	1120.55	5.3929E+00
CR-51	127679.	320.08	2.9647E+02
MN-54	316.	834.83	3.6724E+00
CO-56	335.	846.76	3.8320E+00
CO-58	331.	810.78	3.6794E+00
FE-59	234.	1099.25	7.1705E+00
CO-60	97.	1332.50	3.1208E+00
ZN-65	265.	1115.55	8.5761E+00
SE-75	129006.	264.66	4.3800E+01
KR-85	82378.	514.00	8.4612E+03
SR-85	82371.	514.01	3.8275E+01
Y-88	18.	1836.06	1.8530E+00
Y-91	190.	1204.67	1.3174E+03
NB-94	304.	871.09	3.7480E+00
NB-95	343.	765.78	3.5440E+00
ZR-95	334.	756.73	6.3276E+00
ZRNB-95	460.	724.18	1.7758E+01
RU-103	100676.	497.08	4.3285E+01
RURH-106	24977.	621.93	4.8766E+02
AG-108m	449.	722.94	4.2051E+00
CD-109	100262.	88.03	3.4610E+02
AG-110M	190060.	657.76	7.4134E+01
SN-113	146278.	391.69	5.9606E+01
TE-123m	116955.	159.00	2.0331E+01
SB-124	26262.	602.73	2.4642E+01
SB-125	170520.	427.89	1.5170E+02
TE-125m	107507.	109.27	4.8656E+03
I-129	94120.	39.60	1.8406E+02
I-131	136366.	364.48	4.2748E+01
XE-131m	117755.	163.93	8.9299E+02
BA-133	131998.	356.02	5.3630E+01
CS-134	26145.	604.70	2.4697E+01
CS-136	316.	818.51	3.6284E+00
CS-138	55.	1435.86	1.2695E+01
CE-139	118007.	165.85	2.2016E+01
BA-140	58129.	537.31	1.3228E+02
LA-140	38.	1596.21	2.4710E+00
CE-141	115322.	145.44	3.3512E+01
CE-144	115125.	133.51	1.3920E+02
CEPR-144	115125.	133.51	2.7814E+02
EU-152	114945.	121.78	5.2093E+01
EU-154	125.	1274.51	9.8691E+00
EU-155	100616.	86.54	3.8036E+01
HF-181	131977.	482.18	5.4451E+01
TA-182	94820.	67.75	2.8856E+01
HG-203	127836.	279.20	3.2716E+01
BI-207	33475.	569.70	2.6369E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3009-SAM

Acquisition date : 24-NOV-1998 16:37:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	128111.	277.36	4.2018E+02
PB-210	92214.	46.50	3.1414E+02
BI-212	490.	727.18	6.0214E+01
PB-212	132298.	238.63	5.4859E+01
BI-214	26156.	609.31	5.4195E+01
PB-214	132998.	351.92	4.2100E+02
RA-224	132433.	240.99	6.1437E+02
RA-226	176757.	186.10	6.6106E+02
AC-228	466.	911.21	1.8028E+01
TH-229	100300.	88.47	5.0637E+01
U-232	93235.	57.78	5.9869E+03
PA-233	127254.	312.17	7.5660E+01
UTH-233	131280.	245.34	2.2127E+04
PA-234M	242.	1001.03	5.8322E+02
TH-234	93742.	63.29	3.1220E+02
U-235	173830.	185.71	3.9850E+01
NP-237	100690.	86.48	1.0063E+02
U-237	93066.	59.54	3.4724E+01
NP-238	248.	984.45	1.3873E+01
NP-239	104235.	106.12	5.9460E+01
PU-239	115336.	129.30	2.3739E+05
AM-241	93066.	59.54	3.3164E+01

*****

* 222-S Laboratory Counting Room 25-NOV-1998 13:33:17.90 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26841
 Sample ID: S98T3009-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1137.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:53:57.29 sec
 Dead Time: 14%

Verified by:

11-25-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 24-NOV-1998 19:34:03.36
 Decayed to: 24-NOV-1998 19:34:03.36
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	9.26	34472	1.43	18.65	16	7	2.5			
0	12.46	2401	0.89	25.04	22	6	42.0			
0	21.95	27835	1.29	44.03	41	7	4.6			
0	31.99	901366	1.23	64.10	59	11	0.4			
0	36.51	215833	1.20	73.14	69	8	0.9			
0	75.11*	2387	1.05	150.32	148	7	48.6		AM-243	33.0
0	84.59*	1232	1.26	169.28	167	7	94.0		TH-228	818.
0	187.71	12634	3.91	375.51	371	12	17.6			
0	661.68*	5185140	1.63	1323.46	1315	17	0.1		CS-137	2.114E+05
0	692.34	2587	3.36	1384.78	1378	15	13.1			
0	697.45	224	1.70	1395.00	1393	11	71.6			
0	1322.84	1128	2.03	2646.07	2636	18	8.1			
0	1378.22*	36	1.36	2756.87	2750	14	96.8			

Summary of Nuclide Activity
Sample ID : S98T3009-DUP

Page : 2
Acquisition date : 24-NOV-1998 19:34:03

Total number of lines in spectrum 13
Number of unidentified lines 10
Number of lines tentatively identified by NID 3 23.08%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.114E+05	2.114E+05	0.002E+05	0.09	
TH-228	1.91Y	1.00	8.181E+02	8.181E+02	7.691E+02	94.01	
~~AM-243~~	~~7380.00Y*****~~		~~3.300E+01~~	~~3.300E+01~~	~~1.605E+01~~	~~48.62~~	~~243~~
Total Activity :			2.122E+05	2.122E+05			11-25-98

Grand Total Activity : 2.122E+05 2.122E+05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3009-DUP

Acquisition date : 24-NOV-1998 19:34:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	165098.	477.59	4.6934E+02
NA-24	71.	1368.55	2.9353E+00
K-40	1675.	1460.75	1.2472E+02
AR-41	99.	1293.60	5.0896E+00
SC-46	393.	1120.55	5.2859E+00
CR-51	125741.	320.08	2.9422E+02
MN-54	344.	834.83	3.8270E+00
CO-56	319.	846.76	3.7410E+00
CO-58	308.	810.78	3.5509E+00
FE-59	267.	1099.25	7.6399E+00
CO-60	83.	1332.50	2.8983E+00
ZN-65	268.	1115.55	8.6302E+00
SE-75	127106.	264.66	4.3477E+01
KR-85	81083.	514.00	8.3945E+03
SR-85	81075.	514.01	3.7973E+01
Y-88	20.	1836.06	1.9359E+00
Y-91	173.	1204.67	1.2611E+03
NB-94	332.	871.09	3.9098E+00
NB-95	334.	765.78	3.4960E+00
ZR-95	331.	756.73	6.3070E+00
ZRNB-95	435.	724.18	1.7273E+01
RU-103	98815.	497.08	4.2884E+01
RURH-106	24941.	621.93	4.8731E+02
AG-108m	425.	722.94	4.0924E+00
CD-109	99157.	88.03	3.4419E+02
AG-110M	191953.	657.76	7.4502E+01
SN-113	144022.	391.69	5.9145E+01
TE-123m	115672.	159.00	2.0219E+01
SB-124	26239.	602.73	2.4632E+01
SB-125	168640.	427.89	1.5086E+02
TE-125m	106077.	109.27	4.8332E+03
I-129	92933.	39.60	1.8290E+02
I-131	133470.	364.48	4.2292E+01
XE-131m	116054.	163.93	8.8652E+02
BA-133	131869.	356.02	5.3604E+01
CS-134	25825.	604.70	2.4546E+01
CS-136	310.	818.51	3.5928E+00
CS-138	64.	1435.86	1.3659E+01
CE-139	116478.	165.85	2.1873E+01
BA-140	56804.	537.31	1.3077E+02
LA-140	46.	1596.21	2.6966E+00
CE-141	113938.	145.44	3.3311E+01
CE-144	113386.	133.51	1.3814E+02
CEPR-144	113387.	133.51	2.7604E+02
EU-152	113017.	121.78	5.1655E+01
EU-154	117.	1274.51	9.5631E+00
EU-155	99132.	86.54	3.7755E+01
HF-181	130057.	482.18	5.4054E+01
TA-182	93861.	67.75	2.8710E+01
HG-203	125853.	279.20	3.2462E+01
BI-207	33289.	569.70	2.6296E+01

4310

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3009-DUP

Acquisition date : 24-NOV-1998 19:34:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	125713.	277.36	4.1623E+02
PB-210	90167.	46.50	3.1064E+02
BI-212	477.	727.18	5.9426E+01
PB-212	130773.	238.63	5.4542E+01
BI-214	26083.	609.31	5.4120E+01
PB-214	131287.	351.92	4.1750E+02
RA-224	129837.	240.99	6.0832E+02
RA-226	173680.	186.10	6.5529E+02
AC-228	525.	911.21	1.9116E+01
TH-229	99056.	88.47	5.0323E+01
U-232	91634.	57.78	5.9354E+03
PA-233	125457.	312.17	7.5125E+01
UTH-233	130177.	245.34	2.2034E+04
PA-234M	227.	1001.03	5.6574E+02
TH-234	92822.	63.29	3.1067E+02
U-235	170998.	185.71	3.9525E+01
NP-237	99183.	86.48	9.9879E+01
U-237	92329.	59.54	3.4586E+01
NP-238	231.	984.45	1.3422E+01
NP-239	102802.	106.12	5.9048E+01
PU-239	114166.	129.30	2.3619E+05
AM-241	92329.	59.54	3.3032E+01

LABCORE Completed Worklist Report for Worklist# 26978

Analyst: crj

Instrument: GEA04

Book#: 28B57

Method: LA-548-121 Rev/Mod F-φ

Worklist Comment: BY105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	@GEA-01 CO60-02 SOLID		4.537e-03	4.71e-03	103.813	% Recovery
1 STD	0	0	@GEA-01 CO60-02H SOLID		1	1.59	1.590	% Ct Error
1 STD	0	0	@GEA-01 CS13702 SOLID		7.455e-03	7.84e-03	105.194	% Recovery
1 STD	0	0	@GEA-01 CS13702H SOLID		1	1.49	1.490	% Ct Error
2 BLNK-PREP	0	0	@GEA-01 CO60-02 SOLID		1	<4.68e-3		uCi/g
2 BLNK-PREP	0	0	@GEA-01 CS13702 SOLID		1	<7.60e-3		uCi/g
3 SAMPLE	S98T002598	0 F	@GEA-01 CO60-02 SOLID		N/A	4.599e-03	4.60e-003	uCi/g
3 SAMPLE	S98T002598	0 F	@GEA-01 CO60-02H SOLID		N/A	n/a		% Ct. Error
3 SAMPLE	S98T002598	0 F	@GEA-01 CS13702 SOLID		N/A	3.416e+00		uCi/g
3 SAMPLE	S98T002598	0 F	@GEA-01 CS13702H SOLID		N/A	0.860		% Ct. Error
4 DUF	S98T002598	0 F	@GEA-01 CO60-02 SOLID		<4.60e-3	<4.49e-3		RPD
4 DUF	S98T002598	0 F	@GEA-01 CO60-02H SOLID		1	n/a		% Ct Error
4 DUF	S98T002598	0 F	@GEA-01 CS13702 SOLID		3.42e+00	4.39e+00	24.840	RPD
4 DUF	S98T002598	0 F	@GEA-01 CS13702H SOLID		1.00	0.750	0.750	% Ct Error
5 SAMPLE	S98T002599	0 F	@GEA-01 CO60-02 SOLID		N/A	4.659e-03	4.66e-003	uCi/g
5 SAMPLE	S98T002599	0 F	@GEA-01 CO60-02H SOLID		N/A	n/a		% Ct. Error
5 SAMPLE	S98T002599	0 F	@GEA-01 CS13702 SOLID		N/A	1.024e+01		uCi/g
5 SAMPLE	S98T002599	0 F	@GEA-01 CS13702H SOLID		N/A	0.480		% Ct. Error
6 DUF	S98T002599	0 F	@GEA-01 CO60-02 SOLID		<4.66e-3	<4.45e-3		RPD
6 DUF	S98T002599	0 F	@GEA-01 CO60-02H SOLID		1	n/a		% Ct Error
6 DUF	S98T002599	0 F	@GEA-01 CS13702 SOLID		1.02e+01	1.03e+01	0.976	RPD
6 DUF	S98T002599	0 F	@GEA-01 CS13702H SOLID		1.00	0.480	0.480	% Ct Error
7 SAMPLE	S98T002600	0 F	@GEA-01 CO60-02 SOLID		N/A	5.136e-03	5.14e-003	uCi/g
7 SAMPLE	S98T002600	0 F	@GEA-01 CO60-02H SOLID		N/A	n/a		% Ct. Error
7 SAMPLE	S98T002600	0 F	@GEA-01 CS13702 SOLID		N/A	2.073e+01		uCi/g
7 SAMPLE	S98T002600	0 F	@GEA-01 CS13702H SOLID		N/A	0.340		% Ct. Error
8 DUF	S98T002600	0 F	@GEA-01 CO60-02 SOLID		<5.14e-3	<5.59e-3		RPD
8 DUF	S98T002600	0 F	@GEA-01 CO60-02H SOLID		1	n/a		% Ct Error
8 DUF	S98T002600	0 F	@GEA-01 CS13702 SOLID		2.07e+01	2.18e+01	5.176	RPD
8 DUF	S98T002600	0 F	@GEA-01 CS13702H SOLID		1.00	0.330	0.330	% Ct Error

Comments Section:

Comments for sample# S98T002598 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002599 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002600 and test @GEA-01 .

Units shown for QC (BLK/BKG) may not reflect the actual units.

Sample # S98T002598 reanalyzed on WL# 27168. J Steer
3/4/99

LABCORE Completed Worklist Report for Worklist# 26978

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
		DL=0	=	>	n/a.					

Final page for worklist# 26978

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11-10-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26978

Analyst: CRS Instrument: GEA00 4 Book# 28357

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 FOR @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T002598	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T002598	0	F	@GEA-01	SOLID		
5	SAMPLE	S98T002599	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6	DUP	S98T002599	0	F	@GEA-01	SOLID		
7	SAMPLE	S98T002600	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8	DUP	S98T002600	0	F	@GEA-01	SOLID		

Final page for worklist # 26978

[Signature] 11/18/98
Analyst Signature Date

[Signature] 11/18/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 8-NOV-1998 16:46:04.30 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26978
 Sample ID: WL26978-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4298.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:09.10 sec
 Dead Time: 0.3%

Verified by:

11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 15:55:24.18
 Decayed to: 8-NOV-1998 15:55:24.18
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.57*	23241	1.61	1323.19	1314	18	1.5		CS-137	7.84
0	1173.16	10216	2.00	2346.43	2337	20	2.3		CO-60	4.70
0	1332.43	9276	2.07	2665.01	2655	21	2.2		CO-60	4.73

Summary of Nuclide Activity
Sample ID : WL26978-STD

HNF-1663 REV. 0

Page : 2

Acquisition date : 8-NOV-1998 15:55:24

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.715E+00	4.715E+00	0.075E+00	1.59	
CS-137	30.00Y	1.00	7.844E+00	7.844E+00	0.117E+00	1.49	
Total Activity :			1.256E+01	1.256E+01			

Grand Total Activity :	1.256E+01	1.256E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26978-STD

Acquisition date : 8-NOV-1998 15:55:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1267.	477.59	3.4349E-01
NA-24	48.	1368.55	1.8637E-02
K-40	770.	1460.75	6.8088E-01
AR-41	129.	1293.60	3.2443E-02
SC-46	504.	1120.55	4.7471E-02
CR-51	1225.	320.08	2.3078E-01
MN-54	581.	834.83	4.0256E-02
CO-56	623.	846.76	4.2164E-02
CO-58	552.	810.78	3.8522E-02
FE-59	577.	1099.25	8.8395E-02
ZN-65	549.	1115.55	9.7230E-02
SE-75	1325.	264.66	3.4387E-02
KR-85	922.	514.00	7.5309E+00
SR-85	921.	514.01	3.4035E-02
Y-88	28.	1836.06	1.9005E-02
Y-91	189.	1204.67	1.0422E+01
NB-94	602.	871.09	4.2430E-02
NB-95	559.	765.78	3.6858E-02
ZR-95	565.	756.73	6.7121E-02
ZRNB-95	531.	724.18	1.5693E-01
RU-103	924.	497.08	3.4804E-02
RURH-106	621.	621.93	6.5025E-01
AG-108m	516.	722.94	3.7077E-02
CD-109	1112.	88.03	5.7760E-01
AG-110M	1488.	657.76	5.5059E-02
SN-113	1247.	391.69	4.5048E-02
TE-123m	1246.	159.00	1.6972E-02
SB-124	590.	602.73	3.1256E-02
SB-125	1269.	427.89	1.0828E-01
TE-125m	1202.	109.27	5.6415E+00
I-129	1133.	39.60	1.0929E+01
I-131	1137.	364.48	3.1554E-02
XE-131m	1309.	163.93	7.5088E-01
BA-133	1171.	356.02	4.0840E-02
CS-134	601.	604.70	3.1694E-02
CS-136	550.	818.51	3.8687E-02
CS-138	34.	1435.86	3.4667E-02
CE-139	1348.	165.85	1.8735E-02
BA-140	704.	537.31	1.2267E-01
LA-140	30.	1596.21	1.7681E-02
CE-141	1303.	145.44	3.0035E-02
CE-144	1246.	133.51	1.2931E-01
CEPR-144	1246.	133.51	2.5841E-01
EU-152	1255.	121.78	5.2675E-02
EU-154	110.	1274.51	7.3658E-02
EU-155	1159.	86.54	6.7100E-02
HF-181	989.	482.18	3.9463E-02
TA-182	1103.	67.75	1.0106E-01
HG-203	1340.	279.20	2.6110E-02
BI-207	677.	569.70	3.1703E-02
TL-208	1312.	277.36	3.3145E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26978-STD

Acquisition date : 8-NOV-1998 15:55:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1080.	46.50	6.2843E+00
BI-212	534.	727.18	5.1726E-01
PB-212	1580.	238.63	4.5968E-02
BI-214	765.	609.31	7.8261E-02
PB-214	1280.	351.92	1.3046E-01
RA-224	1492.	240.99	4.9650E-01
RA-226	1598.	186.10	4.8639E-01
AC-228	774.	911.21	1.8689E-01
TH-228	1186.	84.37	1.8535E+00
TH-229	1104.	88.47	8.3306E-02
U-232	1171.	57.78	4.0485E+01
PA-233	1235.	312.17	5.9027E-02
UTH-233	1448.	245.34	1.7869E+01
PA-234M	598.	1001.03	7.2592E+00
TH-234	1049.	63.29	1.3727E+00
U-235	1599.	185.71	2.9588E-02
NP-237	1162.	86.48	1.7793E-01
U-237	1228.	59.54	2.1062E-01
NP-238	615.	984.45	1.7089E-01
NP-239	1183.	106.12	7.1368E-02
PU-239	1215.	129.30	2.2330E+02
AM-241	1228.	59.54	2.0202E-01
AM-243	1098.	74.67	5.1009E-02

*****
 * 222-S Laboratory Counting Room 8-NOV-1998 19:30:41.58 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26978
 Sample ID: WL26978-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SL Uah

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4299.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.37 sec
 Dead Time: 0.1%

Verified by:

W Jones 11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 17:00:02.92
 Decayed to: 8-NOV-1998 17:00:02.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	511.38*	214	2.67	1022.82	1014	23 88.1			

Summary of Nuclide Activity

Sample ID : WL26978-BLK

Acquisition date : 8-NOV-1998 17:00:02

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26978-BLK

Acquisition date : 8-NOV-1998 17:00:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	523.	477.59	7.4252E-02
NA-24	112.	1368.55	9.5443E-03
K-40	2164.	1460.75	3.7728E-01
AR-41	142.	1293.60	1.5032E-02
SC-46	350.	1120.55	1.3260E-02
CR-51	889.	320.08	6.5798E-02
MN-54	271.	834.83	9.2673E-03
CO-56	251.	846.76	9.0403E-03
CO-58	246.	810.78	8.6786E-03
FE-59	211.	1099.25	1.8092E-02
CO-60	132.	1332.50	9.5318E-03
ZN-65	215.	1115.55	2.0574E-02
SE-75	965.	264.66	9.8132E-03
KR-85	712.	514.00	2.2114E+00
SR-85	709.	514.01	9.9870E-03
Y-88	42.	1836.06	7.6376E-03
Y-91	225.	1204.67	3.7821E+00
NB-94	255.	871.09	9.3122E-03
NB-95	321.	765.78	9.3883E-03
ZR-95	303.	756.73	1.6529E-02
ZRNB-95	323.	724.18	4.1069E-02
RU-103	489.	497.08	8.5005E-03
RURH-106	358.	621.93	1.6576E-01
AG-108m	298.	722.94	9.4640E-03
CD-109	927.	88.03	1.7603E-01
AG-110M	339.	657.76	8.9058E-03
SN-113	610.	391.69	1.0581E-02
TE-123m	1139.	159.00	5.4150E-03
SB-124	382.	602.73	8.4365E-03
SB-125	596.	427.89	2.4928E-02
TE-125m	973.	109.27	1.6946E+00
I-129	767.	39.60	3.0091E+00
I-131	670.	364.48	8.1418E-03
XE-131m	1109.	163.93	2.3122E-01
BA-133	685.	356.02	1.0467E-02
CS-134	390.	604.70	8.5563E-03
CS-136	241.	818.51	8.6561E-03
CS-137	841.	661.66	1.5477E-02
CS-138	125.	1435.86	4.3646E-02
CE-139	1044.	165.85	5.5094E-03
BA-140	463.	537.31	3.3383E-02
LA-140	47.	1596.21	7.4074E-03
CE-141	1014.	145.44	8.8561E-03
CE-144	1013.	133.51	3.8953E-02
CEPR-144	1013.	133.51	7.7828E-02
EU-152	1005.	121.78	1.5745E-02
EU-154	168.	1274.51	2.9953E-02
EU-155	1056.	86.54	2.1366E-02
HF-181	471.	482.18	9.1582E-03
TA-182	875.	67.75	3.0073E-02
HG-203	964.	279.20	7.4075E-03

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26978-BLK

Acquisition date : 8-NOV-1998 17:00:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	409.	569.70	8.2675E-03
TL-208	952.	277.36	9.4378E-02
PB-210	770.	46.50	1.7793E+00
BI-212	398.	727.18	1.4951E-01
PB-212	1356.	238.63	1.4215E-02
BI-214	831.	609.31	2.7163E-02
PB-214	1006.	351.92	8.5781E-02
RA-224	1165.	240.99	1.4752E-01
RA-226	1203.	186.10	1.4099E-01
AC-228	519.	911.21	5.1239E-02
TH-228	1147.	84.37	6.0787E-01
TH-229	920.	88.47	2.5393E-02
U-232	773.	57.78	1.1007E+01
PA-233	871.	312.17	1.6584E-02
UTH-233	1028.	245.34	5.0328E+00
PA-234M	241.	1001.03	1.5585E+00
TH-234	819.	63.29	4.0509E-01
U-235	1197.	185.71	8.5525E-03
NP-237	1060.	86.48	5.6700E-02
U-237	797.	59.54	5.6995E-02
NP-238	210.	984.45	3.4225E-02
NP-239	932.	106.12	2.1380E-02
PU-239	1032.	129.30	6.8710E+01
AM-241	797.	59.54	5.4476E-02
AM-243	1105.	74.67	1.7063E-02

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 00:54:43.76 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26978
 Sample ID: S98T2598-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Shuler

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4302.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:14.91 sec
 Dead Time: 0.2%

Verified by:

R. Jones 11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 22:23:51.50
 Decayed to: 8-NOV-1998 22:23:51.50
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.57*	61826	1.63	1323.20	1314	18	0.9		CS-137	6.96

**DATA
NOT USED**

Summary of Nuclide Activity
Sample ID : S98T2598-SAM

Page : 2
Acquisition date : 8-NOV-1998 22:23:51

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	6.956E+00	6.956E+00	0.060E+00	0.86	
Total Activity :			6.956E+00	6.956E+00			

Grand Total Activity : 6.956E+00 6.956E+00

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2598-SAM

Acquisition date : 8-NOV-1998 22:23:51

Nuclide	Bkgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2467.	477.59	1.5907E-01
NA-24	83.	1368.55	8.3108E-03
K-40	2128.	1460.75	3.7414E-01
AR-41	162.	1293.60	1.6000E-02
SC-46	345.	1120.55	1.3166E-02
CR-51	2341.	320.08	1.0597E-01
MN-54	276.	834.83	9.3547E-03
CO-56	255.	846.76	9.1008E-03
CO-58	269.	810.78	9.0628E-03
FE-59	219.	1099.25	1.8419E-02
CO-60	127.	1332.50	9.3697E-03
ZN-65	228.	1115.55	2.1186E-02
SE-75	2518.	264.66	1.5741E-02
KR-85	1535.	514.00	3.2254E+00
SR-85	1533.	514.01	1.4582E-02
Y-88	38.	1836.06	7.2579E-03
Y-91	229.	1204.67	3.8150E+00
NB-94	243.	871.09	9.1135E-03
NB-95	332.	765.78	9.5393E-03
ZR-95	311.	756.73	1.6761E-02
ZRNB-95	315.	724.18	4.0598E-02
RU-103	1679.	497.08	1.5571E-02
RURH-106	720.	621.93	2.3307E-01
AG-108m	289.	722.94	9.3212E-03
CD-109	2191.	88.03	2.6888E-01
AG-110M	3146.	657.76	2.6569E-02
SN-113	2294.	391.69	2.0285E-02
TE-123m	2553.	159.00	8.0608E-03
SB-124	756.	602.73	1.1764E-02
SB-125	2602.	427.89	5.1430E-02
TE-125m	2359.	109.27	2.6216E+00
I-129	2076.	39.60	4.9103E+00
I-131	2248.	364.48	1.4761E-02
XE-131m	2582.	163.93	3.5067E-01
BA-133	2331.	356.02	1.9114E-02
CS-134	760.	604.70	1.1851E-02
CS-136	260.	818.51	8.9800E-03
CS-138	85.	1435.86	3.6406E-02
CE-139	2584.	165.85	8.6103E-03
BA-140	1097.	537.31	5.0931E-02
LA-140	75.	1596.21	9.1803E-03
CE-141	2535.	145.44	1.3914E-02
CE-144	2522.	133.51	6.1038E-02
CEPR-144	2521.	133.51	1.2196E-01
EU-152	2435.	121.78	2.4343E-02
EU-154	152.	1274.51	2.8554E-02
EU-155	2239.	86.54	3.0939E-02
HF-181	2129.	482.18	1.9202E-02
TA-182	2021.	67.75	4.5394E-02
HG-203	2570.	279.20	1.2005E-02
BI-207	847.	569.70	1.1797E-02

DATA
NOT USED

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2598-SAM

Acquisition date : 8-NOV-1998 22:23:51

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2643.	277.36	1.5606E-01
PB-210	2060.	46.50	2.8881E+00
BI-212	370.	727.18	1.4424E-01
PB-212	3103.	238.63	2.1389E-02
BI-214	1229.	609.31	3.2916E-02
PB-214	2695.	351.92	1.3956E-01
RA-224	3023.	240.99	2.3610E-01
RA-226	3285.	186.10	2.3147E-01
AC-228	510.	911.21	5.0810E-02
TH-228	2322.	84.37	8.6048E-01
TH-229	2214.	88.47	3.9125E-02
U-232	1990.	57.78	1.7523E+01
PA-233	2276.	312.17	2.6615E-02
UTH-233	2630.	245.34	7.9951E+00
PA-234M	218.	1001.03	1.4826E+00
TH-234	2128.	63.29	6.4827E-01
U-235	3255.	185.71	1.4010E-02
NP-237	2248.	86.48	8.2129E-02
U-237	2041.	59.54	9.0499E-02
NP-238	256.	984.45	3.7602E-02
NP-239	2298.	106.12	3.3344E-02
PU-239	2552.	129.30	1.0734E+02
AM-241	2041.	59.54	8.6500E-02
AM-243	2421.	74.67	2.5112E-02

DATA
NOT USED

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 03:26:27.73 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26978
 Sample ID: S98T2598-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Sheldon

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4304.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:17.05 sec
 Dead Time: 0.2%

Verified by:

RA Jones 11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 9-NOV-1998 00:55:41.92
 Decayed to: 9-NOV-1998 00:55:41.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.56*	79377	1.62	1323.18	1314	18	0.8		CS-137	8.93
0	727.14*	73	1.05	1454.34	1449	11101.8				
0	1120.51*	77	2.44	2241.13	2235	15100.5				

DATA
NOT USED

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2598-DUP

Acquisition date : 9-NOV-1998 00:55:41

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	8.930E+00	8.930E+00	0.067E+00	0.75	
Total Activity :			8.930E+00	8.930E+00			

Grand Total Activity :	8.930E+00	8.930E+00

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2598-DUP

Acquisition date : 9-NOV-1998 00:55:41

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	3054.	477.59	1.7679E-01
NA-24	113.	1368.55	9.5884E-03
K-40	2116.	1460.75	3.7316E-01
AR-41	129.	1293.60	1.4381E-02
SC-46	385.	1120.55	1.3879E-02
CR-51	2812.	320.08	1.1601E-01
MN-54	295.	834.83	9.6450E-03
CO-56	263.	846.76	9.2499E-03
CO-58	268.	810.78	9.0472E-03
FE-59	238.	1099.25	1.9173E-02
CO-60	121.	1332.50	9.1529E-03
ZN-65	197.	1115.55	1.9726E-02
SE-75	3066.	264.66	1.7350E-02
KR-85	1988.	514.00	3.6637E+00
SR-85	1986.	514.01	1.6565E-02
Y-88	47.	1836.06	8.0262E-03
Y-91	206.	1204.67	3.6219E+00
NB-94	240.	871.09	9.0461E-03
NB-95	323.	765.78	9.4104E-03
ZR-95	270.	756.73	1.5635E-02
ZRNB-95	306.	724.18	4.0021E-02
RU-103	1941.	497.08	1.6728E-02
RURH-106	783.	621.93	2.4283E-01
AG-108m	274.	722.94	9.0869E-03
CD-109	2682.	88.03	2.9716E-01
AG-110M	3917.	657.76	2.9614E-02
SN-113	2798.	391.69	2.2379E-02
TE-123m	2999.	159.00	8.7286E-03
SB-124	898.	602.73	1.2805E-02
SB-125	3131.	427.89	5.6365E-02
TE-125m	2683.	109.27	2.7941E+00
I-129	2394.	39.60	5.2683E+00
I-131	2749.	364.48	1.6307E-02
XE-131m	3081.	163.93	3.8272E-01
BA-133	2692.	356.02	2.0525E-02
CS-134	898.	604.70	1.2862E-02
CS-136	271.	818.51	9.1591E-03
CS-138	81.	1435.86	3.5672E-02
CE-139	3174.	165.85	9.5325E-03
BA-140	1303.	537.31	5.5437E-02
LA-140	61.	1596.21	8.3089E-03
CE-141	2950.	145.44	1.4997E-02
CE-144	3039.	133.51	6.6938E-02
CEPR-144	3039.	133.51	1.3376E-01
EU-152	2901.	121.78	2.6548E-02
EU-154	148.	1274.51	2.8258E-02
EU-155	2757.	86.54	3.4290E-02
HF-181	2581.	482.18	2.1117E-02
TA-182	2467.	67.75	5.0088E-02
HG-203	2885.	279.20	1.2711E-02
BI-207	925.	569.70	1.2315E-02

DATA
NOT USED

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2598-DUP

Acquisition date : 9-NOV-1998 00:55:41

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2993.	277.36	1.6596E-01
PB-210	2331.	46.50	3.0696E+00
BI-212	379.	727.18	1.4594E-01
PB-212	3674.	238.63	2.3253E-02
BI-214	1421.	609.31	3.5347E-02
PB-214	3124.	351.92	1.5015E-01
RA-224	3496.	240.99	2.5373E-01
RA-226	3826.	186.10	2.4961E-01
AC-228	490.	911.21	4.9804E-02
TH-228	2691.	84.37	9.2547E-01
TH-229	2658.	88.47	4.2822E-02
U-232	2547.	57.78	1.9794E+01
PA-233	2797.	312.17	2.9471E-02
UTH-233	3099.	245.34	8.6711E+00
PA-234M	220.	1001.03	1.4893E+00
TH-234	2525.	63.29	7.0539E-01
U-235	3828.	185.71	1.5182E-02
NP-237	2758.	86.48	9.0857E-02
U-237	2467.	59.54	9.9381E-02
NP-238	188.	984.45	3.2458E-02
NP-239	2825.	106.12	3.6926E-02
PU-239	2934.	129.30	1.1500E+02
AM-241	2467.	59.54	9.4987E-02
AM-243	2772.	74.67	2.6849E-02

DATA
NOT USED

*****

* 222-S Laboratory Counting Room 9-NOV-1998 06:10:11.03 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26978
 Sample ID: S98T2599-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4305.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:30.16 sec
 Dead Time: 0.3%

Verified by:

RAJmer 11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 9-NOV-1998 03:39:11.02
 Decayed to: 9-NOV-1998 03:39:11.02
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	32.51	285	1.32	65.17	63	6	81.2			
0	661.55*	187755	1.64	1323.16	1314	18	0.5		CS-137	21.1
0	1662.40	35	4.36	3325.05	3319	13	78.2			

Summary of Nuclide Activity
 Sample ID : S98T2599-SAM

Page : 2
 Acquisition date : 9-NOV-1998 03:39:11

Total number of lines in spectrum	3	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	1	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.112E+01	2.112E+01	0.010E+01	0.48	
Total Activity :			2.112E+01	2.112E+01			

Grand Total Activity : 2.112E+01 2.112E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2599-SAM

Acquisition date : 9-NOV-1998 03:39:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	6510.	477.59	2.5728E-01
NA-24	116.	1368.55	9.7087E-03
K-40	2206.	1460.75	3.8086E-01
AR-41	141.	1293.60	1.4994E-02
SC-46	354.	1120.55	1.3336E-02
CR-51	5452.	320.08	1.6106E-01
MN-54	250.	834.83	8.9049E-03
CO-56	292.	846.76	9.7212E-03
CO-58	245.	810.78	8.6681E-03
FE-59	194.	1099.25	1.7370E-02
CO-60	134.	1332.50	9.6015E-03
ZN-65	220.	1115.55	2.0802E-02
SE-75	5839.	264.66	2.3874E-02
KR-85	3678.	514.00	4.9671E+00
SR-85	3676.	514.01	2.2459E-02
Y-88	47.	1836.06	8.0044E-03
Y-91	215.	1204.67	3.7000E+00
NB-94	266.	871.09	9.5054E-03
NB-95	313.	765.78	9.2694E-03
ZR-95	364.	756.73	1.8088E-02
ZRNB-95	311.	724.18	4.0352E-02
RU-103	4091.	497.08	2.4186E-02
RURH-106	1503.	621.93	3.3447E-01
AG-108m	287.	722.94	9.3019E-03
CD-109	4888.	88.03	3.9999E-01
AG-110M	8946.	657.76	4.4615E-02
SN-113	5844.	391.69	3.2234E-02
TE-123m	5521.	159.00	1.1811E-02
SB-124	1574.	602.73	1.6869E-02
SB-125	6548.	427.89	8.1253E-02
TE-125m	5093.	109.27	3.8379E+00
I-129	4554.	39.60	7.2422E+00
I-131	5391.	364.48	2.2763E-02
XE-131m	5626.	163.93	5.1574E-01
BA-133	5404.	356.02	2.8984E-02
CS-134	1545.	604.70	1.6790E-02
CS-136	257.	818.51	8.9266E-03
CS-138	110.	1435.86	4.1155E-02
CE-139	5611.	165.85	1.2641E-02
BA-140	2588.	537.31	7.7750E-02
LA-140	68.	1596.21	8.7503E-03
CE-141	5516.	145.44	2.0446E-02
CE-144	5540.	133.51	9.0135E-02
CEPR-144	5540.	133.51	1.8012E-01
EU-152	5335.	121.78	3.5899E-02
EU-154	151.	1274.51	2.8464E-02
EU-155	4902.	86.54	4.5594E-02
HF-181	5271.	482.18	3.0077E-02
TA-182	4682.	67.75	6.8788E-02
HG-203	5646.	279.20	1.7729E-02
BI-207	1680.	569.70	1.6520E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2599-SAM

Acquisition date : 9-NOV-1998 03:39:11

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	5666.	277.36	2.2769E-01
PB-210	4541.	46.50	4.2700E+00
BI-212	387.	727.18	1.4744E-01
PB-212	6460.	238.63	3.0763E-02
BI-214	2052.	609.31	4.2372E-02
PB-214	5939.	351.92	2.0673E-01
RA-224	6367.	240.99	3.4156E-01
RA-226	7581.	186.10	3.5043E-01
AC-228	531.	911.21	5.1812E-02
TH-228	4998.	84.37	1.2576E+00
TH-229	4876.	88.47	5.7834E-02
U-232	4530.	57.78	2.6325E+01
PA-233	5455.	312.17	4.1028E-02
UTH-233	6229.	245.34	1.2256E+01
PA-234M	259.	1001.03	1.6129E+00
TH-234	4611.	63.29	9.5040E-01
U-235	7511.	185.71	2.1210E-02
NP-237	4908.	86.48	1.2086E-01
U-237	4656.	59.54	1.3609E-01
NP-238	228.	984.45	3.5563E-02
NP-239	4939.	106.12	4.8702E-02
PU-239	5458.	129.30	1.5639E+02
AM-241	4656.	59.54	1.3008E-01
AM-243	4911.	74.67	3.5639E-02

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 09:14:04.87 *
 *****
 >>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26978
 Sample ID: S98T002599-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<<
 Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4306.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:30.05 sec
 Dead Time: 0.3%

Verified by:

11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<
 Sample Count Time: 9-NOV-1998 06:42:58.63
 Decayed to: 9-NOV-1998 06:42:58.63
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<
 Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60
 *****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
5	26.20	2041	0.91	52.56	49	32	7.4	2.88E+01		
5	27.16*	2305	1.24	54.46	49	32	9.3			
5	28.27*	2343	1.45	56.68	49	32	11.4			
5	29.50*	1903	1.41	59.16	49	32	13.0			
5	30.70*	1380	1.28	61.56	49	32	17.5			
5	31.97	1439	1.50	64.09	49	32	17.8			
5	33.27	1240	1.44	66.68	49	32	20.4			
5	34.81	1043	1.52	69.77	49	32	24.1			
5	36.40	667	1.52	72.94	49	32	37.3			
0	661.55*	187456	1.63	1323.16	1314	18	0.5		CS-137	21.1
0	1379.15	139	8.27	2758.46	2745	27	51.5			
0	1590.66	94	5.81	3181.55	3172	18	44.6			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002599-DUP

Acquisition date : 9-NOV-1998 06:42:58

Total number of lines in spectrum 12
 Number of unidentified lines 10
 Number of lines tentatively identified by NID 2 16.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~1.012E+02~~	~~1.012E+02~~	~~0.132E+02~~	~~13.00~~	~~RM~~
CS-137	30.00Y	1.00	2.109E+01	2.109E+01	0.010E+01	0.48	
Total Activity :			1.223E+02	1.223E+02			11-10-98

Grand Total Activity : 1.223E+02 1.223E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002599-DUP

Acquisition date : 9-NOV-1998 06:42:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	6346.	477.59	2.5403E-01
NA-24	106.	1368.55	9.3101E-03
K-40	2300.	1460.75	3.8884E-01
AR-41	134.	1293.60	1.4649E-02
SC-46	363.	1120.55	1.3485E-02
CR-51	5409.	320.08	1.6041E-01
MN-54	290.	834.83	9.5733E-03
CO-56	240.	846.76	8.8405E-03
CO-58	235.	810.78	8.4903E-03
FE-59	194.	1099.25	1.7385E-02
CO-60	121.	1332.50	9.1564E-03
ZN-65	224.	1115.55	2.1013E-02
SE-75	5869.	264.66	2.3935E-02
KR-85	3516.	514.00	4.8572E+00
SR-85	3514.	514.01	2.1965E-02
Y-88	38.	1836.06	7.2268E-03
Y-91	243.	1204.67	3.9226E+00
NB-94	261.	871.09	9.4243E-03
NB-95	299.	765.78	9.0656E-03
ZR-95	283.	756.73	1.6014E-02
ZRNB-95	329.	724.18	4.1462E-02
RU-103	4020.	497.08	2.3977E-02
RURH-106	1477.	621.93	3.3161E-01
AG-108m	308.	722.94	9.6128E-03
CD-109	4822.	88.03	3.9729E-01
AG-110M	8926.	657.76	4.4565E-02
SN-113	5903.	391.69	3.2395E-02
TE-123m	5574.	159.00	1.1866E-02
SB-124	1575.	602.73	1.6873E-02
SB-125	6607.	427.89	8.1612E-02
TE-125m	5066.	109.27	3.8276E+00
I-131	5408.	364.48	2.2800E-02
XE-131m	5574.	163.93	5.1341E-01
BA-133	5509.	356.02	2.9262E-02
CS-134	1584.	604.70	1.6998E-02
CS-136	242.	818.51	8.6824E-03
CS-138	100.	1435.86	3.9284E-02
CE-139	5659.	165.85	1.2695E-02
BA-140	2474.	537.31	7.6047E-02
LA-140	62.	1596.21	8.4022E-03
CE-141	5608.	145.44	2.0615E-02
CE-144	5521.	133.51	8.9987E-02
CEPR-144	5522.	133.51	1.7982E-01
EU-152	5390.	121.78	3.6083E-02
EU-154	146.	1274.51	2.7998E-02
EU-155	4808.	86.54	4.5163E-02
HF-181	5265.	482.18	3.0062E-02
TA-182	4571.	67.75	6.7980E-02
HG-203	5775.	279.20	1.7929E-02
BI-207	1720.	569.70	1.6712E-02
TL-208	5779.	277.36	2.2994E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002599-DUP

Acquisition date : 9-NOV-1998 06:42:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	4435.	46.50	4.2202E+00
BI-212	379.	727.18	1.4600E-01
PB-212	6732.	238.63	3.1398E-02
BI-214	2010.	609.31	4.1939E-02
PB-214	5879.	351.92	2.0568E-01
RA-224	6364.	240.99	3.4146E-01
RA-226	7422.	186.10	3.4676E-01
AC-228	509.	911.21	5.0722E-02
TH-228	4829.	84.37	1.2363E+00
TH-229	4817.	88.47	5.7488E-02
U-232	4428.	57.78	2.6030E+01
PA-233	5462.	312.17	4.1055E-02
UTH-233	6151.	245.34	1.2180E+01
PA-234M	207.	1001.03	1.4456E+00
TH-234	4664.	63.29	9.5571E-01
U-235	7359.	185.71	2.0996E-02
NP-237	4806.	86.48	1.1961E-01
U-237	4481.	59.54	1.3353E-01
NP-238	215.	984.45	3.4580E-02
NP-239	4796.	106.12	4.7995E-02
PU-239	5497.	129.30	1.5696E+02
AM-241	4481.	59.54	1.2763E-01
AM-243	4935.	74.67	3.5725E-02

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 12:00:56.94 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26978
 Sample ID: S98T2600-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4307.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:52.75 sec
 Dead Time: 0.6%

Verified by:

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 09:29:34.44
 Decayed to: 9-NOV-1998 09:29:34.44
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
7	26.47	5400	1.22	53.08	49	33	4.8	9.97E+01		
7	28.05*	5702	1.80	56.25	49	33	7.1			
7	29.50*	3937	1.73	59.16	49	33	9.8			
7	30.97	3138	1.81	62.08	49	33	13.0		I-129	209.
7	32.19	2250	1.46	64.53	49	33	15.4			
7	33.51	2033	1.47	67.18	49	33	16.9			
7	34.99	1689	1.53	70.13	49	33	20.2			
7	36.49	1509	1.65	73.14	49	33	24.4			
7	38.63	705	1.36	77.40	49	33	43.9		I-129	16.1
0	661.54*	371407	1.63	1323.14	1314	18	0.3		CS-137	41.8
0	1591.06	96	6.00	3182.34	3171	22	56.8			
0	1796.29	18	1.29	3592.89	3584	11	122.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2600-SAM

Acquisition date : 9-NOV-1998 09:29:34

Total number of lines in spectrum	12	
Number of unidentified lines	9	
Number of lines tentatively identified by NID	3	25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected	Decay Corr			
			uCi/L	uCi/L	2-Sigma Error	%Error	
I-129	5.73E+09D	1.00	1.613E+01	1.613E+01	0.709E+01	43.94	
CS-137	30.00Y	1.00	4.178E+01	4.178E+01	0.014E+01	0.34	
Total Activity :			5.792E+01	5.792E+01			

Grand Total Activity :	5.792E+01	5.792E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2600-SAM

Acquisition date : 9-NOV-1998 09:29:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	12527.	477.59	3.5617E-01
NA-24	105.	1368.55	9.2524E-03
K-40	2126.	1460.75	3.7403E-01
AR-41	164.	1293.60	1.6137E-02
SC-46	381.	1120.55	1.3815E-02
CR-51	10073.	320.08	2.1846E-01
MN-54	289.	834.83	9.5596E-03
CO-56	263.	846.76	9.2535E-03
CO-58	259.	810.78	8.8986E-03
FE-59	228.	1099.25	1.8800E-02
CO-60	157.	1332.50	1.0356E-02
ZN-65	192.	1115.55	1.9495E-02
SE-75	10681.	264.66	3.2224E-02
KR-85	6529.	514.00	6.6018E+00
SR-85	6527.	514.01	2.9857E-02
Y-88	54.	1836.06	8.5358E-03
Y-91	238.	1204.67	3.8809E+00
NB-94	247.	871.09	9.1699E-03
NB-95	329.	765.78	9.4971E-03
ZR-95	317.	756.73	1.6898E-02
ZRNB-95	370.	724.18	4.3872E-02
RU-103	7550.	497.08	3.2778E-02
RURH-106	2695.	621.93	4.4622E-01
AG-108m	343.	722.94	1.0131E-02
CD-109	9187.	88.03	5.4715E-01
AG-110M	17720.	657.76	6.2679E-02
SN-113	10951.	391.69	4.4034E-02
TE-123m	10297.	159.00	1.6095E-02
SB-124	2779.	602.73	2.2336E-02
SB-125	12733.	427.89	1.1307E-01
TE-125m	9432.	109.27	5.2115E+00
I-131	10169.	364.48	3.1198E-02
XE-131m	10477.	163.93	7.0239E-01
BA-133	10060.	356.02	3.9463E-02
CS-134	2637.	604.70	2.1861E-02
CS-136	287.	818.51	9.4306E-03
CS-138	84.	1435.86	3.6301E-02
CE-139	10385.	165.85	1.7163E-02
BA-140	4534.	537.31	1.0263E-01
LA-140	72.	1596.21	8.9836E-03
CE-141	10181.	145.44	2.7722E-02
CE-144	10141.	133.51	1.2170E-01
CEPR-144	10141.	133.51	2.4319E-01
EU-152	10181.	121.78	4.9483E-02
EU-154	197.	1274.51	3.2370E-02
EU-155	9354.	86.54	6.2843E-02
HF-181	9945.	482.18	4.1226E-02
TA-182	8452.	67.75	9.2226E-02
HG-203	10421.	279.20	2.4038E-02
BI-207	2997.	569.70	2.1984E-02
TL-208	10400.	277.36	3.0786E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2600-SAM

Acquisition date : 9-NOV-1998 09:29:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	8419.	46.50	5.8010E+00
BI-212	379.	727.18	1.4592E-01
PB-212	12015.	238.63	4.1873E-02
BI-214	3102.	609.31	5.1971E-02
PB-214	10465.	351.92	2.7454E-01
RA-224	11654.	240.99	4.6122E-01
RA-226	13745.	186.10	4.7105E-01
AC-228	484.	911.21	4.9535E-02
TH-228	9188.	84.37	1.7015E+00
TH-229	9161.	88.47	7.9096E-02
U-232	8464.	57.78	3.5902E+01
PA-233	10153.	312.17	5.5856E-02
UTH-233	11310.	245.34	1.6484E+01
PA-234M	242.	1001.03	1.5587E+00
TH-234	8591.	63.29	1.2943E+00
U-235	13645.	185.71	2.8539E-02
NP-237	9360.	86.48	1.6653E-01
U-237	8496.	59.54	1.8344E-01
NP-238	214.	984.45	3.4555E-02
NP-239	9310.	106.12	6.6716E-02
PU-239	10204.	129.30	2.1339E+02
AM-241	8496.	59.54	1.7533E-01
AM-243	9169.	74.67	4.8589E-02

*****

* 222-S Laboratory Counting Room 9-NOV-1998 15:11:01.44 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26978
 Sample ID: S98T2600-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4309.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:54.89 sec
 Dead Time: 0.6%

Verified by:

11-10-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 12:39:37.10
 Decayed to: 9-NOV-1998 12:39:37.10
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.53*	389653	1.64	1323.13	1314	18	0.3		CS-137	43.8
0	911.12*	108	1.94	1822.31	1814	18	92.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2600-DUP

Acquisition date : 9-NOV-1998 12:39:37

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlfe	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
CS-137	30.00Y	1.00	4.384E+01	4.384E+01	0.015E+01	0.33			
			-----	-----					
Total Activity :			4.384E+01	4.384E+01					

Grand Total Activity :	4.384E+01	4.384E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2600-DUP

Acquisition date : 9-NOV-1998 12:39:37

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	12766.	477.59	3.5954E-01
NA-24	91.	1368.55	8.6410E-03
K-40	2179.	1460.75	3.7856E-01
AR-41	141.	1293.60	1.5027E-02
SC-46	351.	1120.55	1.3277E-02
CR-51	10434.	320.08	2.2231E-01
MN-54	298.	834.83	9.7005E-03
CO-56	267.	846.76	9.3076E-03
CO-58	269.	810.78	9.0599E-03
FE-59	208.	1099.25	1.7980E-02
CO-60	186.	1332.50	1.1221E-02
ZN-65	236.	1115.55	2.1543E-02
SE-75	11305.	264.66	3.3148E-02
KR-85	6729.	514.00	6.7012E+00
SR-85	6725.	514.01	3.0304E-02
Y-88	37.	1836.06	7.2088E-03
Y-91	224.	1204.67	3.7755E+00
NB-94	283.	871.09	9.7974E-03
NB-95	315.	765.78	9.2994E-03
ZR-95	347.	756.73	1.7657E-02
ZRNB-95	370.	724.18	4.3886E-02
RU-103	8158.	497.08	3.4064E-02
RURH-106	2736.	621.93	4.4955E-01
AG-108m	339.	722.94	1.0077E-02
CD-109	9523.	88.03	5.5700E-01
AG-110M	18648.	657.76	6.4293E-02
SN-113	11506.	391.69	4.5130E-02
TE-123m	10776.	159.00	1.6463E-02
SB-124	2837.	602.73	2.2564E-02
SB-125	13226.	427.89	1.1523E-01
TE-125m	9920.	109.27	5.3439E+00
I-129	9032.	39.60	1.0173E+01
I-131	10834.	364.48	3.2196E-02
XE-131m	10798.	163.93	7.1300E-01
BA-133	10531.	356.02	4.0371E-02
CS-134	2851.	604.70	2.2723E-02
CS-136	233.	818.51	8.5140E-03
CS-138	104.	1435.86	4.0262E-02
CE-139	11012.	165.85	1.7670E-02
BA-140	4781.	537.31	1.0537E-01
LA-140	87.	1596.21	9.8268E-03
CE-141	10786.	145.44	2.8529E-02
CE-144	10755.	133.51	1.2532E-01
CEPR-144	10754.	133.51	2.5040E-01
EU-152	10558.	121.78	5.0387E-02
EU-154	186.	1274.51	3.1505E-02
EU-155	9659.	86.54	6.3853E-02
HF-181	10423.	482.18	4.2198E-02
TA-182	9089.	67.75	9.5618E-02
HG-203	10873.	279.20	2.4551E-02
BI-207	3171.	569.70	2.2607E-02

Minimum Detectable Activity Report (continued)

Sample ID : S98T2600-DUP

Acquisition date : 9-NOV-1998 12:39:37

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	10848.	277.36	3.1439E-01
PB-210	8787.	46.50	5.9256E+00
BI-212	404.	727.18	1.5053E-01
PB-212	12386.	238.63	4.2511E-02
BI-214	3180.	609.31	5.2613E-02
PB-214	10917.	351.92	2.8044E-01
RA-224	12080.	240.99	4.6953E-01
RA-226	14352.	186.10	4.8128E-01
AC-228	546.	911.21	5.2517E-02
TH-228	9749.	84.37	1.7523E+00
TH-229	9474.	88.47	8.0429E-02
U-232	8919.	57.78	3.6848E+01
PA-233	10617.	312.17	5.7113E-02
UTH-233	11797.	245.34	1.6834E+01
PA-234M	244.	1001.03	1.5658E+00
TH-234	9012.	63.29	1.3254E+00
U-235	14210.	185.71	2.9120E-02
NP-237	9666.	86.48	1.6922E-01
U-237	8751.	59.54	1.8615E-01
NP-238	196.	984.45	3.3058E-02
NP-239	9931.	106.12	6.8892E-02
PU-239	10601.	129.30	2.1748E+02
AM-241	8750.	59.54	1.7792E-01
AM-243	9688.	74.67	4.9937E-02

LABCORE Completed Worklist Report for Worklist# 26979

Analyst: glt

Instrument: GEA01

Book#: 28B57

Method: LA-548-121

Rev/Mod F-φ

Worklist Comment: BY105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	GEA-01	CO60-02	SOLID	4.537e-03	4.93e-03	108.662 % Recovery
1 STD	0	0	GEA-01	CO60-02E	SOLID	1	3.09	3.090 % Ct Error
1 STD	0	0	GEA-01	CS13702	SOLID	7.455e-03	7.47e-03	100.201 % Recovery
1 STD	0	0	GEA-01	CS13702E	SOLID	1	2.92	2.920 % Ct Error
2 BLNK-PREP	0	0	GEA-01	CO60-02	SOLID	1	<2.99e-5	uCi/g
3 SAMPLE	S98T002601	0 F	GEA-01	CO60-02	SOLID	N/A	< 1.328e-02	1.33e-002 uCi/g
3 SAMPLE	S98T002601	0 F	GEA-01	CO60-02E	SOLID	N/A	n/a	% Ct. Error
3 SAMPLE	S98T002601	0 F	GEA-01	CS13702	SOLID	N/A	8.769e+00	uCi/g
3 SAMPLE	S98T002601	0 F	GEA-01	CS13702E	SOLID	N/A	1.00	% Ct. Error
4 DUP	S98T002601	0 F	GEA-01	CO60-02	SOLID	<1.33e-2	<1.30e-2	RPD
4 DUP	S98T002601	0 F	GEA-01	CO60-02E	SOLID	1	n/a	% Ct Error
4 DUP	S98T002601	0 F	GEA-01	CS13702	SOLID	8.77e+00	8.83e+00	0.682 RPD
4 DUP	S98T002601	0 F	GEA-01	CS13702E	SOLID	1.00	0.990	% Ct Error
5 SAMPLE	S98T002602	0 F	GEA-01	CO60-02	SOLID	N/A	< 5.135e+00	5.135 uCi/g
5 SAMPLE	S98T002602	0 F	GEA-01	CO60-02E	SOLID	N/A	n/a	% Ct. Error
5 SAMPLE	S98T002602	0 F	GEA-01	CS13702	SOLID	N/A	2.604e+02	uCi/g
5 SAMPLE	S98T002602	0 F	GEA-01	CS13702E	SOLID	N/A	4.31	% Ct. Error
6 DUP	S98T002602	0 F	GEA-01	CO60-02	SOLID	<5.13e0	<5.12e0	RPD
6 DUP	S98T002602	0 F	GEA-01	CO60-02E	SOLID	1	n/a	% Ct Error
6 DUP	S98T002602	0 F	GEA-01	CS13702	SOLID	2.60e+02	2.49e+02	4.322 RPD
6 DUP	S98T002602	0 F	GEA-01	CS13702E	SOLID	1.00	4.50	% Ct Error
7 SAMPLE	S98T002603	0 F	GEA-01	CO60-02	SOLID	N/A	< 5.690e+00	5.690 uCi/g
7 SAMPLE	S98T002603	0 F	GEA-01	CO60-02E	SOLID	N/A	n/a	% Ct. Error
7 SAMPLE	S98T002603	0 F	GEA-01	CS13702	SOLID	N/A	2.172e+02	uCi/g
7 SAMPLE	S98T002603	0 F	GEA-01	CS13702E	SOLID	N/A	4.82	% Ct. Error
8 DUP	S98T002603	0 F	GEA-01	CO60-02	SOLID	<5.69e0	<6.44e0	RPD
8 DUP	S98T002603	0 F	GEA-01	CO60-02E	SOLID	1	n/a	% Ct Error
8 DUP	S98T002603	0 F	GEA-01	CS13702	SOLID	2.17e+02	2.20e+02	1.373 RPD
8 DUP	S98T002603	0 F	GEA-01	CS13702E	SOLID	1.00	5.04	% Ct Error

Final page for worklist# 26979

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

LABCORE Data Entry Template for Worklist# 26979

Analyst: gll Instrument: GEA00 / Book# 28357

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 FOR @GEA-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	SOLID		
2 BLNK-PREP			@GEA-01	SOLID		
3 SAMPLE	S98T002601 0 F	@GEA-01	SOLID		98000436	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
4 DUP	S98T002601 0 F	@GEA-01	SOLID			
5 SAMPLE	S98T002602 0 F	@GEA-01	SOLID		98000436	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
6 DUP	S98T002602 0 F	@GEA-01	SOLID			
7 SAMPLE	S98T002603 0 F	@GEA-01	SOLID		98000436	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
8 DUP	S98T002603 0 F	@GEA-01	SOLID			

Final page for worklist # 26979

Ag LPLT 11/8/98
Analyst Signature Date

Ag LPLT 11-7-98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 8-NOV-1998 11:35:11.45 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26979
 Sample ID: WL26979-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Revised

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1020.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:14.56 sec
 Dead Time: 0.5%

Verified by:

RA Jones H9-98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 10:44:24.77
 Decayed to: 8-NOV-1998 10:44:24.77
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	31.95	927	0.97	64.02	60	10	15.2			
0	36.40	369	1.17	72.91	69	10	31.1			
0	59.24	464	1.07	118.60	113	11	28.3			
									U-232	58.0
									AM-241	0.321
0	238.50*	93	0.88	477.08	473	8	96.8			
0	661.75*	6171	1.42	1323.60	1316	15	2.9		CS-137	7.47
0	1173.22	2904	1.76	2346.74	2337	20	4.4		CO-60	5.00
0	1332.47	2550	1.81	2665.34	2656	18	4.3		CO-60	4.86

Summary of Nuclide Activity

Page : 2

Sample ID : WL26979-STD

Acquisition date : 8-NOV-1998 10:44:24

Total number of lines in spectrum	7	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	5	71.43%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.929E+00	4.929E+00	0.152E+00	3.09	
CS-137	30.00Y	1.00	7.472E+00	7.472E+00	0.218E+00	2.92	
~~U-232~~	~~68.90Y~~	~~1.00~~	~~5.798E+01~~	~~5.798E+01~~	~~1.643E+01~~	~~28.34~~	~~28.34~~ <i>raj</i>
AM-241	432.20Y	1.00	3.215E-01	3.215E-01	0.911E-01	28.34	
Total Activity :			7.070E+01	7.070E+01			11-9-98

Grand Total Activity : 7.070E+01 7.070E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26979-STD

Acquisition date : 8-NOV-1998 10:44:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	419.	477.59	7.2078E-01
NA-24	23.	1368.55	4.9807E-02
K-40	549.	1460.75	2.1426E+00
AR-41	42.	1293.60	7.1988E-02
SC-46	182.	1120.55	1.0824E-01
CR-51	438.	320.08	5.2859E-01
MN-54	196.	834.83	8.6634E-02
CO-56	173.	846.76	8.2596E-02
CO-58	174.	810.78	8.0069E-02
FE-59	165.	1099.25	1.8005E-01
ZN-65	164.	1115.55	2.0228E-01
SE-75	527.	264.66	8.5137E-02
KR-85	362.	514.00	1.7128E+01
SR-85	361.	514.01	7.7395E-02
Y-88	10.	1836.06	4.1889E-02
Y-91	76.	1204.67	2.5398E+01
NB-94	197.	871.09	9.0291E-02
NB-95	156.	765.78	7.2054E-02
ZR-95	192.	756.73	1.4387E-01
ZRNB-95	182.	724.18	3.3709E-01
RU-103	323.	497.08	7.5029E-02
RURH-106	228.	621.93	1.4323E+00
AG-108m	181.	722.94	8.0581E-02
CD-109	611.	88.03	8.1983E-01
AG-110M	317.	657.76	9.2666E-02
SN-113	435.	391.69	9.9055E-02
TE-123m	684.	159.00	4.7128E-02
SB-124	196.	602.73	6.5515E-02
SB-125	426.	427.89	2.3109E-01
TE-125m	656.	109.27	1.1525E+01
I-129	430.	39.60	3.7901E-01
I-131	381.	364.48	6.8729E-02
XE-131m	662.	163.93	2.0259E+00
BA-133	426.	356.02	9.2927E-02
CS-134	202.	604.70	6.6952E-02
CS-136	207.	818.51	8.7581E-02
CS-138	18.	1435.86	9.8323E-02
CE-139	659.	165.85	4.9891E-02
BA-140	256.	537.31	2.6918E-01
LA-140	15.	1596.21	4.7895E-02
CE-141	696.	145.44	7.8836E-02
CE-144	726.	133.51	3.3475E-01
CEPR-144	726.	133.51	6.6887E-01
EU-152	720.	121.78	1.2492E-01
EU-154	46.	1274.51	1.8267E-01
EU-155	580.	86.54	8.7667E-02
HF-181	374.	482.18	8.8532E-02
TA-182	554.	67.75	6.6998E-02
HG-203	519.	279.20	6.3334E-02
BI-207	227.	569.70	6.6813E-02
TL-208	531.	277.36	8.2224E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26979-STD

Acquisition date : 8-NOV-1998 10:44:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	509.	46.50	7.0685E-01
BI-212	215.	727.18	1.2000E+00
PB-212	692.	238.63	1.2025E-01
BI-214	317.	609.31	1.8246E-01
PB-214	496.	351.92	3.0662E-01
RA-224	647.	240.99	1.2918E+00
RA-226	739.	186.10	1.2949E+00
AC-228	318.	911.21	4.4498E-01
TH-228	626.	84.37	2.3488E+00
TH-229	613.	88.47	1.2017E-01
PA-233	450.	312.17	1.3699E-01
UTH-233	558.	245.34	4.3840E+01
PA-234M	191.	1001.03	1.5472E+01
TH-234	572.	63.29	7.4060E-01
U-235	738.	185.71	7.8652E-02
NP-237	583.	86.48	2.3249E-01
U-237	919.	59.54	1.0380E-01
NP-238	203.	984.45	3.6910E-01
NP-239	640.	106.12	1.3955E-01
PU-239	721.	129.30	5.6866E+02
AM-243	651.	74.67	4.9850E-02

*****
 * 222-S Laboratory Counting Room 8-NOV-1998 14:25:39.83 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26979
 Sample ID: WL#26979-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

S. Chuan

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1021.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:32.34 sec
 Dead Time: 0.4%

Verified by:

R. Jones 11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 11:54:36.10
 Decayed to: 8-NOV-1998 11:54:36.10
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	13.03	176	0.99	26.18	23	7	67.4			
0	16.34	280	1.25	32.79	29	8	47.0			
0	62.99*	132	0.81	126.10	123	7	95.2			
0	661.44*	138	1.56	1322.98	1317	12	55.3		CS-137	5.577E-02
0	795.32*	58	1.82	1590.78	1587	9	80.2			
0	1593.92	95	5.91	3188.44	3173	39	60.3			

Summary of Nuclide Activity

Page : 2

Sample ID : WL#26979-BLK

Acquisition date : 8-NOV-1998 11:54:36

Total number of lines in spectrum	6	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	3	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~5.577E-02~~	~~5.577E-02~~	~~3.082E-02~~	~~55.26~~			~~11-9-98~~
Total Activity :			5.577E-02	5.577E-02					

Grand Total Activity : 5.577E-02   5.577E-02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL#26979-BLK

Acquisition date : 8-NOV-1998 11:54:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	339.	477.59	2.1695E-01
NA-24	42.	1368.55	2.2639E-02
K-40	1688.	1460.75	1.2396E+00
AR-41	104.	1293.60	4.8562E-02
SC-46	260.	1120.55	4.2873E-02
CR-51	595.	320.08	2.0483E-01
MN-54	170.	834.83	2.6983E-02
CO-56	150.	846.76	2.5796E-02
CO-58	157.	810.78	2.5458E-02
FE-59	114.	1099.25	5.0233E-02
CO-60	91.	1332.50	2.9862E-02
ZN-65	125.	1115.55	5.9270E-02
SE-75	774.	264.66	3.4242E-02
KR-85	367.	514.00	5.7534E+00
SR-85	367.	514.01	2.6001E-02
Y-88	14.	1836.06	1.6207E-02
Y-91	120.	1204.67	1.0495E+01
NB-94	158.	871.09	2.7062E-02
NB-95	191.	765.78	2.6413E-02
ZR-95	184.	756.73	4.7061E-02
ZRNB-95	194.	724.18	1.1568E-01
RU-103	318.	497.08	2.4829E-02
RURH-106	204.	621.93	4.5196E-01
AG-108m	183.	722.94	2.6978E-02
CD-109	1211.	88.03	3.8219E-01
AG-110M	199.	657.76	2.4668E-02
SN-113	417.	391.69	3.2353E-02
TE-123m	1374.	159.00	2.2122E-02
SB-124	257.	602.73	2.4907E-02
SB-125	378.	427.89	7.2693E-02
TE-125m	1321.	109.27	5.4166E+00
I-129	708.	39.60	1.6120E-01
I-131	494.	364.48	2.6090E-02
XE-131m	1348.	163.93	9.5872E-01
BA-133	499.	356.02	3.3454E-02
CS-134	267.	604.70	2.5504E-02
CS-136	189.	818.51	2.8062E-02
CS-138	55.	1435.86	1.1087E-01
CE-139	1317.	165.85	2.3356E-02
BA-140	298.	537.31	9.6689E-02
LA-140	48.	1596.21	2.7244E-02
CE-141	1392.	145.44	3.6956E-02
CE-144	1524.	133.51	1.6065E-01
CEPR-144	1524.	133.51	3.2097E-01
EU-152	1471.	121.78	5.9129E-02
EU-154	94.	1274.51	8.5181E-02
EU-155	1231.	86.54	4.2267E-02
HF-181	331.	482.18	2.7813E-02
TA-182	971.	67.75	2.9402E-02
HG-203	759.	279.20	2.5446E-02
BI-207	266.	569.70	2.4047E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL#26979-BLK

Acquisition date : 8-NOV-1998 11:54:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	781.	277.36	3.3101E-01
PB-210	854.	46.50	3.0450E-01
BI-212	270.	727.18	4.4684E-01
PB-212	1248.	238.63	5.3536E-02
BI-214	641.	609.31	8.5633E-02
PB-214	807.	351.92	2.8837E-01
RA-224	1016.	240.99	5.4079E-01
RA-226	1214.	186.10	5.5070E-01
AC-228	335.	911.21	1.5203E-01
TH-228	1294.	84.37	1.1175E+00
TH-229	1201.	88.47	5.5673E-02
U-232	994.	57.78	6.2229E+00
PA-233	627.	312.17	5.3701E-02
UTH-233	835.	245.34	1.7795E+01
PA-234M	140.	1001.03	4.4462E+00
TH-234	1131.	63.29	3.4480E-01
U-235	1225.	185.71	3.3629E-02
NP-237	1242.	86.48	1.1229E-01
U-237	1010.	59.54	3.6366E-02
NP-238	126.	984.45	9.9194E-02
NP-239	1327.	106.12	6.7205E-02
PU-239	1482.	129.30	2.7000E+02
AM-241	1010.	59.54	3.4761E-02
AM-243	1357.	74.67	2.3824E-02

*****

* 222-S Laboratory Counting Room 8-NOV-1998 18:13:22.79 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26979
 Sample ID: S98T002601-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Skidale

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1023.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:43.49 sec
 Dead Time: 0.5%

Verified by:

R Jones 11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 8-NOV-1998 15:42:08.51
 Decayed to: 8-NOV-1998 15:42:08.51
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	16.22	349	1.07	32.57	29	8	51.3			
0	22.00	172	1.10	44.13	42	6	82.1			
0	31.99	7727	1.06	64.11	60	10	4.1			
0	36.43	2281	1.07	72.98	69	9	9.7			
0	661.75*	44854	1.45	1323.60	1316	15	1.0		CS-137	18.1

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002601-SAM

Acquisition date : 8-NOV-1998 15:42:08

Total number of lines in spectrum	5	
Number of unidentified lines	4	
Number of lines tentatively identified by NID	1	20.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.810E+01	1.810E+01	0.018E+01	1.00	
Total Activity :			1.810E+01	1.810E+01			

Grand Total Activity :	1.810E+01	1.810E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002601-SAM

Acquisition date : 8-NOV-1998 15:42:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1677.	477.59	4.7428E-01
NA-24	68.	1368.55	2.8296E-02
K-40	1764.	1460.75	1.2668E+00
AR-41	75.	1293.60	4.1716E-02
SC-46	259.	1120.55	4.2796E-02
CR-51	1673.	320.08	3.4012E-01
MN-54	164.	834.83	2.6531E-02
CO-56	162.	846.76	2.6714E-02
CO-58	135.	810.78	2.3700E-02
FE-59	135.	1099.25	5.4515E-02
CO-60	76.	1332.50	2.7412E-02
ZN-65	129.	1115.55	6.0108E-02
SE-75	1871.	264.66	5.2845E-02
KR-85	1055.	514.00	9.6306E+00
SR-85	1054.	514.01	4.3544E-02
Y-88	30.	1836.06	2.2727E-02
Y-91	154.	1204.67	1.1788E+01
NB-94	137.	871.09	2.5308E-02
NB-95	186.	765.78	2.6129E-02
ZR-95	200.	756.73	4.8901E-02
ZRNB-95	234.	724.18	1.2673E-01
RU-103	1156.	497.08	4.6621E-02
RURH-106	405.	621.93	6.3014E-01
AG-108m	233.	722.94	3.0278E-02
CD-109	2041.	88.03	4.9435E-01
AG-110M	1315.	657.76	6.1954E-02
SN-113	1719.	391.69	6.4763E-02
TE-123m	2418.	159.00	2.9236E-02
SB-124	430.	602.73	3.1963E-02
SB-125	1917.	427.89	1.6116E-01
TE-125m	2272.	109.27	7.0757E+00
I-129	1539.	39.60	2.3607E-01
I-131	1681.	364.48	4.7551E-02
XE-131m	2308.	163.93	1.2500E+00
BA-133	1631.	356.02	5.9770E-02
CS-134	445.	604.70	3.2650E-02
CS-136	154.	818.51	2.5399E-02
CS-138	49.	1435.86	1.0482E-01
CE-139	2288.	165.85	3.0668E-02
BA-140	750.	537.31	1.5154E-01
LA-140	32.	1596.21	2.2649E-02
CE-141	2440.	145.44	4.8739E-02
CE-144	2461.	133.51	2.0353E-01
CEPR-144	2461.	133.51	4.0667E-01
EU-152	2413.	121.78	7.5482E-02
EU-154	101.	1274.51	8.8250E-02
EU-155	2138.	86.54	5.5482E-02
HF-181	1395.	482.18	5.6191E-02
TA-182	1762.	67.75	3.9410E-02
HG-203	1803.	279.20	3.8932E-02
BI-207	560.	569.70	3.4503E-02

4359

Minimum Detectable Activity Report (continued)

Sample ID : S98T002601-SAM

Acquisition date : 8-NOV-1998 15:42:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1836.	277.36	5.0391E-01
PB-210	1674.	46.50	4.2388E-01
BI-212	232.	727.18	4.1519E-01
PB-212	2253.	238.63	7.1633E-02
BI-214	818.	609.31	9.6485E-02
PB-214	1963.	351.92	4.4714E-01
RA-224	2053.	240.99	7.6465E-01
RA-226	2861.	186.10	8.4065E-01
AC-228	379.	911.21	1.6146E-01
TH-228	2111.	84.37	1.4224E+00
TH-229	2007.	88.47	7.1715E-02
U-232	1830.	57.78	8.4018E+00
PA-233	1730.	312.17	8.8402E-02
UTH-233	2001.	245.34	2.7352E+01
PA-234M	152.	1001.03	4.6243E+00
TH-234	1872.	63.29	4.4186E-01
U-235	2867.	185.71	5.1148E-02
NP-237	2149.	86.48	1.4711E-01
U-237	1788.	59.54	4.8181E-02
NP-238	151.	984.45	1.0807E-01
NP-239	2238.	106.12	8.6961E-02
PU-239	2492.	129.30	3.4891E+02
AM-241	1788.	59.54	4.6053E-02
AM-243	2267.	74.67	3.0683E-02

```

*****
* 222-S Laboratory Counting Room    9-NOV-1998 00:46:01.90
*****
>>>>>>>> SAMPLE INFORMATION <<<<<<<<<
Worklist #:          26979
Sample ID:           S98T2601-DUP
Sample Size:         1.00000E-03 L
Dilution Factor:    1.00000E+00
Removed by:          SLU/du

>>>>>>>> COUNT INFORMATION <<<<<<<<<
Detector ID:         GEA1
File Number:         dka300:[spec.GEA1]1g1026.cnf
Geometry:            41
Count Time:          0 02:30:00.00 sec
Real Time:           0 02:30:43.66 sec
Dead Time:           0.5%
Verified by:         RL Jones 11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<
Sample Count Time:   8-NOV-1998 22:14:48.98
Decayed to:          8-NOV-1998 22:14:48.98
Standard Deviations: 2
Analysis Library:     ENVGEA
Analyst:             SLH2
Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<
Date of last energy calibration: 23-SEP-1998 14:57:30.35
Date of last efficiency calibration: 20-MAR-1996 09:10:52.48
*****

```

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	22.17	359	1.12	44.46	41	8	48.4			
0	31.98	7960	1.03	64.09	60	9	3.7			
0	36.47	2198	0.98	73.06	69	9	9.9			
0	661.76*	44992	1.43	1323.62	1316	15	1.0		CS-137	18.2
0	1764.52*	34	2.69	3529.80	3523	14	94.3			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2601-DUP

Acquisition date : 8-NOV-1998 22:14:48

Total number of lines in spectrum	5	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	2	40.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.816E+01	1.816E+01	0.018E+01	0.99	
Total Activity :			1.816E+01	1.816E+01			

Grand Total Activity :	1.816E+01	1.816E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2601-DUP

Acquisition date : 8-NOV-1998 22:14:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1763.	477.59	4.8605E-01
NA-24	49.	1368.55	2.4325E-02
K-40	1775.	1460.75	1.2707E+00
AR-41	84.	1293.60	4.3989E-02
SC-46	231.	1120.55	4.0494E-02
CR-51	1664.	320.08	3.3923E-01
MN-54	161.	834.83	2.6275E-02
CO-56	152.	846.76	2.5928E-02
CO-58	158.	810.78	2.5509E-02
FE-59	125.	1099.25	5.2604E-02
CO-60	72.	1332.50	2.6817E-02
ZN-65	140.	1115.55	6.2607E-02
SE-75	1846.	264.66	5.2493E-02
KR-85	1080.	514.00	9.7419E+00
SR-85	1079.	514.01	4.4044E-02
Y-88	23.	1836.06	2.0319E-02
Y-91	159.	1204.67	1.1990E+01
NB-94	165.	871.09	2.7648E-02
NB-95	207.	765.78	2.7451E-02
ZR-95	201.	756.73	4.9103E-02
ZRNB-95	220.	724.18	1.2304E-01
RU-103	1110.	497.08	4.5692E-02
RURH-106	424.	621.93	6.4464E-01
AG-108m	204.	722.94	2.8400E-02
CD-109	2052.	88.03	4.9564E-01
AG-110M	1295.	657.76	6.1474E-02
SN-113	1733.	391.69	6.5037E-02
TE-123m	2332.	159.00	2.8714E-02
SB-124	466.	602.73	3.3270E-02
SB-125	1797.	427.89	1.5608E-01
TE-125m	2353.	109.27	7.2002E+00
I-129	1533.	39.60	2.3556E-01
I-131	1657.	364.48	4.7210E-02
XE-131m	2240.	163.93	1.2317E+00
BA-133	1630.	356.02	5.9766E-02
CS-134	458.	604.70	3.3123E-02
CS-136	153.	818.51	2.5301E-02
CS-138	57.	1435.86	1.1252E-01
CE-139	2251.	165.85	3.0422E-02
BA-140	767.	537.31	1.5312E-01
LA-140	39.	1596.21	2.4773E-02
CE-141	2390.	145.44	4.8247E-02
CE-144	2355.	133.51	1.9912E-01
CEPR-144	2355.	133.51	3.9791E-01
EU-152	2500.	121.78	7.6816E-02
EU-154	113.	1274.51	9.2994E-02
EU-155	2112.	86.54	5.5157E-02
HF-181	1390.	482.18	5.6089E-02
TA-182	1916.	67.75	4.1074E-02
HG-203	1811.	279.20	3.9016E-02
BI-207	528.	569.70	3.3511E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2601-DUP

Acquisition date : 8-NOV-1998 22:14:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1898.	277.36	5.1223E-01
PB-210	1702.	46.50	4.2744E-01
BI-212	253.	727.18	4.3230E-01
PB-212	2330.	238.63	7.2838E-02
BI-214	853.	609.31	9.8493E-02
PB-214	1966.	351.92	4.4745E-01
RA-224	2084.	240.99	7.7040E-01
RA-226	2722.	186.10	8.2014E-01
AC-228	328.	911.21	1.5050E-01
TH-228	2181.	84.37	1.4455E+00
TH-229	2082.	88.47	7.3021E-02
U-232	1856.	57.78	8.4596E+00
PA-233	1762.	312.17	8.9212E-02
UTH-233	2019.	245.34	2.7472E+01
PA-234M	143.	1001.03	4.4932E+00
TH-234	1914.	63.29	4.4671E-01
U-235	2717.	185.71	4.9809E-02
NP-237	2121.	86.48	1.4618E-01
U-237	1847.	59.54	4.8949E-02
NP-238	123.	984.45	9.7813E-02
NP-239	2148.	106.12	8.5217E-02
PU-239	2402.	129.30	3.4264E+02
AM-241	1847.	59.54	4.6789E-02
AM-243	2190.	74.67	3.0162E-02

*****

* 222-S Laboratory Counting Room 9-NOV-1998 03:20:28.62 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26979
 Sample ID: S98T2602-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.01000E+02

Removed by:

Shade

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEAL
 File Number: dka300:[spec.GEAL]1g1027.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:33.57 sec
 Dead Time: 0.4%

Verified by:

RA Jones 11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 00:49:24.56
 Decayed to: 9-NOV-1998 00:49:24.56
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEAL]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	13.17	208	1.20	26.47	23	7	58.8			
0	16.23	240	1.20	32.58	30	7	50.6			
0	31.98	587	1.02	64.08	60	8	24.1			
0	36.39	173	0.96	72.90	70	7	68.3			
0	661.76*	3326	1.45	1323.61	1316	14	4.3		CS-137	542.

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2602-SAM

Acquisition date : 9-NOV-1998 00:49:24

Total number of lines in spectrum	5	
Number of unidentified lines	4	
Number of lines tentatively identified by NID	1	20.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	5.422E+02	5.422E+02	0.233E+02	4.31	
Total Activity :			5.422E+02	5.422E+02			

Grand Total Activity :	5.422E+02	5.422E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2602-SAM

Acquisition date : 9-NOV-1998 00:49:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	385.	477.59	9.3229E+01
NA-24	53.	1368.55	1.0123E+01
K-40	1728.	1460.75	5.0660E+02
AR-41	98.	1293.60	1.9053E+01
SC-46	264.	1120.55	1.7447E+01
CR-51	656.	320.08	8.6795E+01
MN-54	178.	834.83	1.1148E+01
CO-56	147.	846.76	1.0317E+01
CO-58	189.	810.78	1.1237E+01
FE-59	146.	1099.25	2.2847E+01
CO-60	70.	1332.50	1.0689E+01
ZN-65	113.	1115.55	2.2785E+01
SE-75	879.	264.66	1.4721E+01
KR-85	444.	514.00	2.5482E+03
SR-85	443.	514.01	1.1514E+01
Y-88	24.	1836.06	8.3069E+00
Y-91	150.	1204.67	4.7120E+03
NB-94	162.	871.09	1.1083E+01
NB-95	193.	765.78	1.0743E+01
ZR-95	199.	756.73	1.9701E+01
ZRNB-95	199.	724.18	4.7402E+01
RU-103	379.	497.08	1.0921E+01
RURH-106	229.	621.93	1.9315E+02
AG-108m	212.	722.94	1.1681E+01
CD-109	1309.	88.03	1.6042E+02
AG-110M	264.	657.76	1.1427E+01
SN-113	535.	391.69	1.4754E+01
TE-123m	1379.	159.00	8.9534E+00
SB-124	278.	602.73	1.0460E+01
SB-125	488.	427.89	3.3253E+01
TE-125m	1442.	109.27	2.2843E+03
I-129	838.	39.60	7.0726E+01
I-131	597.	364.48	1.1556E+01
XE-131m	1387.	163.93	3.9280E+02
BA-133	615.	356.02	1.4965E+01
CS-134	261.	604.70	1.0185E+01
CS-136	152.	818.51	1.0218E+01
CS-138	54.	1435.86	4.4341E+01
CE-139	1364.	165.85	9.5995E+00
BA-140	311.	537.31	3.9873E+01
LA-140	38.	1596.21	9.9474E+00
CE-141	1502.	145.44	1.5498E+01
CE-144	1461.	133.51	6.3566E+01
CEPR-144	1461.	133.51	1.2702E+02
EU-152	1520.	121.78	2.4276E+01
EU-154	88.	1274.51	3.3341E+01
EU-155	1358.	86.54	1.7924E+01
HF-181	418.	482.18	1.2580E+01
TA-182	1128.	67.75	1.2783E+01
HG-203	836.	279.20	1.0776E+01
BI-207	280.	569.70	9.9540E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2602-SAM

Acquisition date : 9-NOV-1998 00:49:24

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	810.	277.36	1.3613E+02
PB-210	1027.	46.50	1.3465E+02
BI-212	232.	727.18	1.6777E+02
PB-212	1330.	238.63	2.2316E+01
BI-214	607.	609.31	3.3682E+01
PB-214	905.	351.92	1.2325E+02
RA-224	1108.	240.99	2.2803E+02
RA-226	1386.	186.10	2.3745E+02
AC-228	388.	911.21	6.5985E+01
TH-228	1373.	84.37	4.6482E+02
TH-229	1297.	88.47	2.3362E+01
U-232	1164.	57.78	2.7158E+03
PA-233	669.	312.17	2.2389E+01
UTH-233	936.	245.34	7.6018E+03
PA-234M	149.	1001.03	1.8489E+03
TH-234	1179.	63.29	1.4217E+02
U-235	1381.	185.71	1.4409E+01
NP-237	1362.	86.48	4.7474E+01
U-237	1198.	59.54	1.5977E+01
NP-238	131.	984.45	4.0754E+01
NP-239	1456.	106.12	2.8420E+01
PU-239	1489.	129.30	1.0935E+05
AM-241	1198.	59.54	1.5272E+01
AM-243	1331.	74.67	9.5335E+00

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 06:05:53.10 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26979
 Sample ID: S98T2602-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1028.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:33.04 sec
 Dead Time: 0.4%

Verified by:

11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 03:34:49.85
 Decayed to: 9-NOV-1998 03:34:49.85
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	13.05	185	1.30	26.22	23	7	65.2			
0	16.25	185	1.07	32.62	30	7	65.4			
0	32.06	520	1.07	64.23	60	9	29.1			
0	36.53	212	0.94	73.17	69	8	61.3			
0	661.79*	3193	1.41	1323.69	1317	14	4.5		CS-137	521.
0	2019.98	19	4.50	4040.99	4034	14	73.6			

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2602-DUP

Acquisition date : 9-NOV-1998 03:34:49

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	421.	477.59	9.7292E+01
NA-24	64.	1368.55	1.1074E+01
K-40	1763.	1460.75	5.1164E+02
AR-41	89.	1293.60	1.8199E+01
SC-46	264.	1120.55	1.7458E+01
CR-51	650.	320.08	8.6374E+01
MN-54	151.	834.83	1.0307E+01
CO-56	119.	846.76	9.3123E+00
CO-58	168.	810.78	1.0613E+01
FE-59	141.	1099.25	2.2452E+01
CO-60	71.	1332.50	1.0733E+01
ZN-65	128.	1115.55	2.4182E+01
SE-75	882.	264.66	1.4742E+01
KR-85	413.	514.00	2.4591E+03
SR-85	412.	514.01	1.1111E+01
Y-88	26.	1836.06	8.6690E+00
Y-91	132.	1204.67	4.4377E+03
NB-94	139.	871.09	1.0283E+01
NB-95	161.	765.78	9.8395E+00
ZR-95	170.	756.73	1.8307E+01
ZRNB-95	210.	724.18	4.8585E+01
RU-103	340.	497.08	1.0358E+01
RURH-106	271.	621.93	2.0966E+02
AG-108m	201.	722.94	1.1389E+01
CD-109	1308.	88.03	1.6036E+02
AG-110M	317.	657.76	1.2487E+01
SN-113	546.	391.69	1.4909E+01
TE-123m	1460.	159.00	9.2091E+00
SB-124	258.	602.73	1.0092E+01
SB-125	477.	427.89	3.2884E+01
TE-125m	1438.	109.27	2.2817E+03
I-129	861.	39.60	7.1663E+01
I-131	561.	364.48	1.1209E+01
XE-131m	1380.	163.93	3.9191E+02
BA-133	625.	356.02	1.5079E+01
CS-134	253.	604.70	1.0043E+01
CS-136	168.	818.51	1.0702E+01
CS-138	49.	1435.86	4.2369E+01
CE-139	1424.	165.85	9.8075E+00
BA-140	307.	537.31	3.9642E+01
LA-140	38.	1596.21	9.9509E+00
CE-141	1482.	145.44	1.5397E+01
CE-144	1567.	133.51	6.5809E+01
CEPR-144	1567.	133.51	1.3148E+02
EU-152	1502.	121.78	2.4135E+01
EU-154	98.	1274.51	3.5139E+01
EU-155	1382.	86.54	1.8075E+01
HF-181	378.	482.18	1.1977E+01
TA-182	1054.	67.75	1.2361E+01
HG-203	830.	279.20	1.0741E+01
BI-207	272.	569.70	9.8040E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2602-DUP

Acquisition date : 9-NOV-1998 03:34:49

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	860.	277.36	1.4019E+02
PB-210	1032.	46.50	1.3496E+02
BI-212	250.	727.18	1.7381E+02
PB-212	1326.	238.63	2.2285E+01
BI-214	646.	609.31	3.4724E+01
PB-214	885.	351.92	1.2194E+02
RA-224	1089.	240.99	2.2608E+02
RA-226	1357.	186.10	2.3500E+02
AC-228	386.	911.21	6.5855E+01
TH-228	1417.	84.37	4.7217E+02
TH-229	1311.	88.47	2.3488E+01
U-232	1180.	57.78	2.7346E+03
PA-233	699.	312.17	2.2876E+01
UTH-233	1014.	245.34	7.9079E+03
PA-234M	135.	1001.03	1.7661E+03
TH-234	1206.	63.29	1.4376E+02
U-235	1365.	185.71	1.4330E+01
NP-237	1391.	86.48	4.7967E+01
U-237	1161.	59.54	1.5738E+01
NP-238	156.	984.45	4.4336E+01
NP-239	1393.	106.12	2.7809E+01
PU-239	1551.	129.30	1.1154E+05
AM-241	1161.	59.54	1.5043E+01
AM-243	1481.	74.67	1.0048E+01

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 09:22:36.81 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26979
 Sample ID: S98T2603-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: NISAP-2.cnf + 161029.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:33.64 sec
 Dead Time: 0.4%

Verified by:

11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 06:33:02.97
 Decayed to: 9-NOV-1998 06:33:02.97
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	13.02	199	1.03	26.17	23	7	62.6			
0	16.00	206	1.21	32.13	30	6	52.7			
0	31.98	423	1.01	64.07	60	9	35.2			
0	661.79*	2661	1.30	1323.68	1317	13	4.8		CS-137	434.
0	860.30*	54	2.90	1720.76	1715		12104.8			
0	1460.79*	71	1.90	2922.06	2912	21	46.6		K-40	184.
0	1764.68*	21	1.72	3530.12	3523		16153.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2603-SAM

Acquisition date : 9-NOV-1998 06:33:02

Total number of lines in spectrum	7	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	4	57.14%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected uCi/L	Decay Corr uCi/L			
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.836E+02~~	~~1.836E+02~~	~~2.691E+02~~	~~146.57~~	~~20.8~~
CS-137	30.00Y	1.00	4.339E+02	4.339E+02	0.209E+02	4.82	11.99%
Total Activity :			6.175E+02	6.175E+02			

Grand Total Activity :	6.175E+02	6.175E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2603-SAM

Acquisition date : 9-NOV-1998 06:33:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	411.	477.59	9.6173E+01
NA-24	55.	1368.55	1.0334E+01
AR-41	95.	1293.60	1.8793E+01
SC-46	250.	1120.55	1.6992E+01
CR-51	666.	320.08	8.7420E+01
MN-54	147.	834.83	1.0176E+01
CO-56	133.	846.76	9.8235E+00
CO-58	162.	810.78	1.0426E+01
FE-59	129.	1099.25	2.1551E+01
CO-60	80.	1332.50	1.1368E+01
ZN-65	113.	1115.55	2.2871E+01
SE-75	870.	264.66	1.4644E+01
KR-85	409.	514.00	2.4475E+03
SR-85	408.	514.01	1.1060E+01
Y-88	19.	1836.06	7.5215E+00
Y-91	148.	1204.67	4.6779E+03
NB-94	156.	871.09	1.0861E+01
NB-95	167.	765.78	1.0005E+01
ZR-95	184.	756.73	1.9006E+01
ZRNB-95	213.	724.18	4.8890E+01
RU-103	352.	497.08	1.0539E+01
RURH-106	233.	621.93	1.9488E+02
AG-108m	210.	722.94	1.1651E+01
CD-109	1299.	88.03	1.5984E+02
AG-110M	294.	657.76	1.2036E+01
SN-113	528.	391.69	1.4664E+01
TE-123m	1459.	159.00	9.2044E+00
SB-124	231.	602.73	9.5685E+00
SB-125	390.	427.89	2.9843E+01
TE-125m	1512.	109.27	2.3383E+03
I-129	836.	39.60	7.0632E+01
I-131	532.	364.48	1.0925E+01
XE-131m	1426.	163.93	3.9820E+02
BA-133	583.	356.02	1.4576E+01
CS-134	259.	604.70	1.0142E+01
CS-136	181.	818.51	1.1095E+01
CS-138	46.	1435.86	4.1352E+01
CE-139	1412.	165.85	9.7636E+00
BA-140	269.	537.31	3.7139E+01
LA-140	41.	1596.21	1.0243E+01
CE-141	1505.	145.44	1.5513E+01
CE-144	1518.	133.51	6.4786E+01
CEPR-144	1519.	133.51	1.2947E+02
EU-152	1459.	121.78	2.3794E+01
EU-154	91.	1274.51	3.3940E+01
EU-155	1360.	86.54	1.7934E+01
HF-181	368.	482.18	1.1827E+01
TA-182	1093.	67.75	1.2588E+01
HG-203	811.	279.20	1.0617E+01
BI-207	298.	569.70	1.0248E+01
TL-208	892.	277.36	1.4276E+02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2603-SAM

Acquisition date : 9-NOV-1998 06:33:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1047.	46.50	1.3594E+02
BI-212	241.	727.18	1.7079E+02
PB-212	1295.	238.63	2.2029E+01
BI-214	607.	609.31	3.3692E+01
PB-214	858.	351.92	1.2010E+02
RA-224	1073.	240.99	2.2443E+02
RA-226	1331.	186.10	2.3278E+02
AC-228	354.	911.21	6.3106E+01
TH-228	1396.	84.37	4.6865E+02
TH-229	1271.	88.47	2.3126E+01
U-232	1136.	57.78	2.6834E+03
PA-233	668.	312.17	2.2383E+01
UTH-233	967.	245.34	7.7255E+03
PA-234M	125.	1001.03	1.7023E+03
TH-234	1211.	63.29	1.4402E+02
U-235	1329.	185.71	1.4138E+01
NP-237	1367.	86.48	4.7544E+01
U-237	1209.	59.54	1.6049E+01
NP-238	133.	984.45	4.1118E+01
NP-239	1366.	106.12	2.7544E+01
PU-239	1426.	129.30	1.0702E+05
AM-241	1209.	59.54	1.5340E+01
AM-243	1427.	74.67	9.8645E+00

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 11:56:29.83 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26979
 Sample ID: S98T2603-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.01000E+02

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1030.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:38.91 sec
 Dead Time: 0.4%

Verified by:

11-9-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 9-NOV-1998 09:25:18.29
 Decayed to: 9-NOV-1998 09:25:18.29
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	14.86	314	3.98	29.84	22	14	72.6			
0	32.10	452	1.04	64.32	60	9	32.7			
0	661.75*	2651	1.39	1323.61	1316	14	5.0		CS-137	432.
0	911.11*	76	1.64	1822.40	1815	16	96.0			
0	1377.72*	35	2.53	2755.88	2749	13	84.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2603-DUP

Acquisition date : 9-NOV-1998 09:25:18

Total number of lines in spectrum	5	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	2	40.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	4.322E+02	4.322E+02	0.218E+02	5.04	
Total Activity :			4.322E+02	4.322E+02			

Grand Total Activity :	4.322E+02	4.322E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2603-DUP

Acquisition date : 9-NOV-1998 09:25:18

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	405.	477.59	9.5553E+01
NA-24	54.	1368.55	1.0281E+01
K-40	1704.	1460.75	5.0312E+02
AR-41	84.	1293.60	1.7755E+01
SC-46	256.	1120.55	1.7173E+01
CR-51	652.	320.08	8.6498E+01
MN-54	161.	834.83	1.0609E+01
CO-56	153.	846.76	1.0507E+01
CO-58	151.	810.78	1.0093E+01
FE-59	115.	1099.25	2.0389E+01
CO-60	100.	1332.50	1.2649E+01
ZN-65	148.	1115.55	2.5911E+01
SE-75	827.	264.66	1.4290E+01
KR-85	441.	514.00	2.5397E+03
SR-85	440.	514.01	1.1478E+01
Y-88	19.	1836.06	7.4777E+00
Y-91	148.	1204.67	4.6726E+03
NB-94	180.	871.09	1.1636E+01
NB-95	194.	765.78	1.0751E+01
ZR-95	174.	756.73	1.8488E+01
ZRNB-95	188.	724.18	4.6034E+01
RU-103	360.	497.08	1.0651E+01
RURH-106	251.	621.93	2.0179E+02
AG-108m	184.	722.94	1.0926E+01
CD-109	1283.	88.03	1.5887E+02
AG-110M	284.	657.76	1.1843E+01
SN-113	518.	391.69	1.4534E+01
TE-123m	1377.	159.00	8.9463E+00
SB-124	248.	602.73	9.8881E+00
SB-125	480.	427.89	3.3007E+01
TE-125m	1433.	109.27	2.2778E+03
I-129	884.	39.60	7.2594E+01
I-131	543.	364.48	1.1029E+01
XE-131m	1398.	163.93	3.9443E+02
BA-133	579.	356.02	1.4524E+01
CS-134	242.	604.70	9.8244E+00
CS-136	161.	818.51	1.0474E+01
CS-138	44.	1435.86	4.0598E+01
CE-139	1387.	165.85	9.6807E+00
BA-140	325.	537.31	4.0712E+01
LA-140	40.	1596.21	1.0075E+01
CE-141	1523.	145.44	1.5604E+01
CE-144	1548.	133.51	6.5397E+01
CEPR-144	1547.	133.51	1.3066E+02
EU-152	1482.	121.78	2.3977E+01
EU-154	87.	1274.51	3.3223E+01
EU-155	1346.	86.54	1.7843E+01
HF-181	372.	482.18	1.1895E+01
TA-182	1062.	67.75	1.2411E+01
HG-203	807.	279.20	1.0592E+01
BI-207	312.	569.70	1.0478E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2603-DUP

Acquisition date : 9-NOV-1998 09:25:18

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	793.	277.36	1.3473E+02
PB-210	1014.	46.50	1.3381E+02
BI-212	227.	727.18	1.6580E+02
PB-212	1270.	238.63	2.1814E+01
BI-214	602.	609.31	3.3543E+01
PB-214	919.	351.92	1.2429E+02
RA-224	1115.	240.99	2.2873E+02
RA-226	1427.	186.10	2.4095E+02
AC-228	401.	911.21	6.7050E+01
TH-228	1411.	84.37	4.7119E+02
TH-229	1257.	88.47	2.3005E+01
U-232	1176.	57.78	2.7298E+03
PA-233	680.	312.17	2.2572E+01
UTH-233	955.	245.34	7.6794E+03
PA-234M	144.	1001.03	1.8198E+03
TH-234	1188.	63.29	1.4267E+02
U-235	1429.	185.71	1.4653E+01
NP-237	1354.	86.48	4.7334E+01
U-237	1111.	59.54	1.5398E+01
NP-238	147.	984.45	4.3078E+01
NP-239	1377.	106.12	2.7647E+01
PU-239	1576.	129.30	1.1242E+05
AM-241	1111.	59.54	1.4718E+01
AM-243	1483.	74.67	1.0056E+01

LABCORE Completed Worklist Report for Worklist# 26980

Analyst: crj

Instrument: GEA03

Book# 28057

Method: LA-548-121 Rev/Mod F-φ

Worklist Comment: BY105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@GEA-01	CO60-02	SOLID	4.537e-03	4.92e-03	108.442	% Recovery
1 STD	0	@GEA-01	CO60-02R	SOLID	1	1.61	1.610	% Ct Error
1 STD	0	@GEA-01	CS13702	SOLID	7.455e-03	7.71e-03	103.421	% Recovery
1 STD	0	@GEA-01	CS13702R	SOLID	1	1.54	1.540	% Ct Error
2 BLNK-PREP	0	@GEA-01	CO60-02	SOLID	1	<6.60e-2		uCi/g
2 BLNK-PREP	0	@GEA-01	CS13702	SOLID	1	<1.73e-1		uCi/g
3 SAMPLR	S98T002845	0 F	@GEA-01	CO60-02	SOLID	N/A	8.40e-002	uCi/g
3 SAMPLR	S98T002845	0 F	@GEA-01	CO60-02R	SOLID	N/A		% Ct. Error
3 SAMPLR	S98T002845	0 F	@GEA-01	CS13702	SOLID	N/A	1.065e+02	uCi/g
3 SAMPLR	S98T002845	0 F	@GEA-01	CS13702R	SOLID	N/A	0.860	% Ct. Error
4 DUP	S98T002845	0 F	@GEA-01	CO60-02	SOLID	<8.40e-2	<8.35e-2	RPD
4 DUP	S98T002845	0 F	@GEA-01	CO60-02R	SOLID	1	n/a	% Ct Error
4 DUP	S98T002845	0 F	@GEA-01	CS13702	SOLID	1.06e+02	1.10e+02	3.704 RPD
4 DUP	S98T002845	0 F	@GEA-01	CS13702R	SOLID	1.00	0.850	0.850 % Ct Error
5 SAMPLR	S98T002918	0 F	@GEA-01	CO60-02	SOLID	N/A	<1.723e-02	1.72e-002 uCi/g
5 SAMPLR	S98T002918	0 F	@GEA-01	CO60-02R	SOLID	N/A		% Ct. Error
5 SAMPLR	S98T002918	0 F	@GEA-01	CS13702	SOLID	N/A	2.634e+01	uCi/g
5 SAMPLR	S98T002918	0 F	@GEA-01	CS13702R	SOLID	N/A	0.770	% Ct. Error
6 DUP	S98T002918	0 F	@GEA-01	CO60-02	SOLID	<1.72e-2	<1.36e-2	RPD
6 DUP	S98T002918	0 F	@GEA-01	CO60-02R	SOLID	1	n/a	% Ct Error
6 DUP	S98T002918	0 F	@GEA-01	CS13702	SOLID	2.63e+01	2.17e+01	19.167 RPD
6 DUP	S98T002918	0 F	@GEA-01	CS13702R	SOLID	1.00	0.840	0.840 % Ct Error
7 SAMPLR	S98T002919	0 F	@GEA-01	CO60-02	SOLID	N/A	8.052e-03	8.05e-003 uCi/g
7 SAMPLR	S98T002919	0 F	@GEA-01	CO60-02R	SOLID	N/A		% Ct. Error
7 SAMPLR	S98T002919	0 F	@GEA-01	CS13702	SOLID	N/A	7.655e+00	uCi/g
7 SAMPLR	S98T002919	0 F	@GEA-01	CS13702R	SOLID	N/A	1.02	% Ct. Error
8 DUP	S98T002919	0 F	@GEA-01	CO60-02	SOLID	<8.05e-3	<8.29e-3	RPD
8 DUP	S98T002919	0 F	@GEA-01	CO60-02R	SOLID	1	n/a	% Ct Error
8 DUP	S98T002919	0 F	@GEA-01	CS13702	SOLID	7.67e+00	7.51e+00	2.108 RPD
8 DUP	S98T002919	0 F	@GEA-01	CS13702R	SOLID	1.00	1.04	1.040 % Ct Error

Comments Section:

Comments for sample# S98T002845 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002918 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002919 and test @GEA-01 .

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26980

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit

DL=0 => n/a.

Final page for worklist# 26980

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11-10-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26980

Analyst: CRJ Instrument: GEA00 #3 Book# 28357

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 FOR @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T002845	0	F	@GEA-01	SOLID	98000437	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T002845	0	F	@GEA-01	SOLID		
5	SAMPLE	S98T002918	0	F	@GEA-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6	DUP	S98T002918	0	F	@GEA-01	SOLID		
7	SAMPLE	S98T002919	0	F	@GEA-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8	DUP	S98T002919	0	F	@GEA-01	SOLID		

Final page for worklist # 26980

[Signature] 11/8/98
Analyst Signature Date

[Signature] 11/12/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 8-NOV-1998 18:15:13.16 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: WL#26980-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SLUde

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3825.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:08.38 sec
 Dead Time: 0.1%

Verified by:

John Relyea 10 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 15:44:35.73
 Decayed to: 8-NOV-1998 15:44:35.73
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	59.85	175	1.34	120.40	116	9	86.3		AM-241	0.263
0	661.77*	20474	1.56	1323.72	1316	16	1.5		CS-137	7.71
0	1173.14	9332	1.91	2346.44	2338	19	2.3		CO-60	4.91
0	1332.35	8456	2.03	2664.93	2654	22	2.2		CO-60	4.93

Summary of Nuclide Activity
Sample ID : WL#26980-STD

Page : 2
Acquisition date : 8-NOV-1998 15:44:35

Total number of lines in spectrum 4
Number of unidentified lines 0
Number of lines tentatively identified by NID 4 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.920E+00	4.920E+00	0.079E+00	1.61	
CS-137	30.00Y	1.00	7.714E+00	7.714E+00	0.119E+00	1.54	
~~AM-241~~	~~432.20Y*****~~		~~2.633E-01~~	~~2.633E-01~~	~~2.272E-01~~	~~86.28~~	~~MS~~
Total Activity :			1.290E+01	1.290E+01			11/10/98

Grand Total Activity : 1.290E+01 1.290E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : WL#26980-STD

Page : 3

Acquisition date : 8-NOV-1998 15:44:35

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1078.	477.59	3.6703E-01
NA-24	32.	1368.55	1.8328E-02
K-40	433.	1460.75	5.8327E-01
AR-41	91.	1293.60	4.1760E-02
SC-46	404.	1120.55	4.8591E-02
CR-51	1045.	320.08	2.7213E-01
MN-54	484.	834.83	4.1144E-02
CO-56	534.	846.76	4.3740E-02
CO-58	476.	810.78	4.0054E-02
FE-59	473.	1099.25	9.1293E-02
ZN-65	462.	1115.55	1.0187E-01
SE-75	1241.	264.66	4.4336E-02
KR-85	776.	514.00	7.8950E+00
SR-85	775.	514.01	3.5701E-02
Y-88	19.	1836.06	1.7188E-02
Y-91	158.	1204.67	1.0961E+01
NB-94	525.	871.09	4.4457E-02
NB-95	427.	765.78	3.6044E-02
ZR-95	442.	756.73	6.6453E-02
ZRNB-95	403.	724.18	1.5300E-01
RU-103	780.	497.08	3.6795E-02
RURH-106	482.	621.93	6.4358E-01
AG-108m	403.	722.94	3.6646E-02
CD-109	1104.	88.03	6.4597E-01
AG-110M	846.	657.76	4.6600E-02
SN-113	993.	391.69	4.8653E-02
TE-123m	1292.	159.00	2.1678E-02
SB-124	480.	602.73	3.1729E-02
SB-125	1089.	427.89	1.1879E-01
TE-125m	1162.	109.27	6.5117E+00
I-129	1167.	39.60	0.0000E+00
I-131	991.	364.48	3.6434E-02
XE-131m	1266.	163.93	9.3540E-01
BA-133	1018.	356.02	4.7268E-02
CS-134	507.	604.70	3.2741E-02
CS-136	461.	818.51	3.9730E-02
CS-138	26.	1435.86	7.1522E-02
CE-139	1262.	165.85	2.2980E-02
BA-140	549.	537.31	1.2347E-01
LA-140	16.	1596.21	1.5337E-02
CE-141	1245.	145.44	3.6163E-02
CE-144	1276.	133.51	1.5852E-01
CEPR-144	1276.	133.51	3.1676E-01
EU-152	1134.	121.78	5.9822E-02
EU-154	86.	1274.51	7.4993E-02
EU-155	1098.	86.54	7.3043E-02
HF-181	951.	482.18	4.4739E-02
TA-182	945.	67.75	1.0194E-01
HG-203	1171.	279.20	3.2148E-02
BI-207	530.	569.70	3.1727E-02
TL-208	1163.	277.36	4.1140E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL#26980-STD

Acquisition date : 8-NOV-1998 15:44:35

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1251.	46.50	0.0000E+00
BI-212	398.	727.18	4.9962E-01
PB-212	1431.	238.63	5.9142E-02
BI-214	560.	609.31	7.5381E-02
PB-214	1089.	351.92	3.3286E-01
RA-224	1393.	240.99	6.5232E-01
RA-226	1471.	186.10	6.0770E-01
AC-228	643.	911.21	1.9156E-01
TH-228	1133.	84.37	2.0149E+00
TH-229	1118.	88.47	9.4163E-02
U-232	1313.	57.78	5.7445E+01
PA-233	1066.	312.17	7.0409E-02
UTH-233	1359.	245.34	2.3342E+01
PA-234M	513.	1001.03	7.6130E+00
TH-234	1106.	63.29	1.5879E+00
U-235	1458.	185.71	3.6768E-02
NP-237	1102.	86.48	1.9378E-01
U-237	1394.	59.54	2.7740E-01
NP-238	493.	984.45	1.7504E-01
NP-239	1150.	106.12	8.3029E-02
PU-239	1219.	129.30	2.6951E+02
AM-243	1082.	74.67	5.5025E-02

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 00:49:04.39 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: WL26980-BLNK
 Sample Size: 1.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3826.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:02.10 sec
 Dead Time: 0.0%

Verified by:

John Rehyer 12 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 8-NOV-1998 22:18:27.12
 Decayed to: 8-NOV-1998 22:18:27.12
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	861.33	29	1.25	1722.78	1720	8	65.1			
0	1119.21*	31	4.84	2238.57	2231	15	92.6			

Summary of Nuclide Activity

Page : 2

Sample ID : WL26980-BLNK

Acquisition date : 8-NOV-1998 22:18:27

Total number of lines in spectrum	2
Number of unidentified lines	0
Number of lines tentatively identified by NID	2 100.00%

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL26980-BLNK

Acquisition date : 8-NOV-1998 22:18:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	99.	477.59	1.1581E+00
NA-24	21.	1368.55	1.5261E-01
K-40	383.	1460.75	5.4993E+00
AR-41	37.	1293.60	2.7545E-01
SC-46	67.	1120.55	2.0611E-01
CR-51	166.	320.08	1.1144E+00
MN-54	53.	834.83	1.4325E-01
CO-56	55.	846.76	1.4787E-01
CO-58	41.	810.78	1.2470E-01
FE-59	37.	1099.25	2.7372E-01
CO-60	19.	1332.50	1.3455E-01
ZN-65	39.	1115.55	3.1609E-01
SE-75	187.	264.66	1.7637E-01
KR-85	150.	514.00	3.5621E+01
SR-85	150.	514.01	1.6089E-01
Y-88	7.	1836.06	1.1485E-01
Y-91	40.	1204.67	5.7575E+01
NB-94	38.	871.09	1.2713E-01
NB-95	49.	765.78	1.2916E-01
ZR-95	58.	756.73	2.5225E-01
ZRNB-95	51.	724.18	5.7300E-01
RU-103	88.	497.08	1.2824E-01
RURH-106	69.	621.93	2.5390E+00
AG-108m	55.	722.94	1.4146E-01
CD-109	210.	88.03	2.8780E+00
AG-110M	61.	657.76	1.3203E-01
SN-113	137.	391.69	1.8629E-01
TE-123m	251.	159.00	9.7401E-02
SB-124	82.	602.73	1.3583E-01
SB-125	115.	427.89	3.9943E-01
TE-125m	243.	109.27	3.0384E+01
I-129	360.	39.60	0.0000E+00
I-131	130.	364.48	1.3609E-01
XE-131m	238.	163.93	4.1434E+00
BA-133	131.	356.02	1.7487E-01
CS-134	83.	604.70	1.3715E-01
CS-136	58.	818.51	1.4780E-01
CS-137	383.	661.66	3.5282E-01
CS-138	30.	1435.86	7.6240E-01
CE-139	217.	165.85	9.7453E-02
BA-140	74.	537.31	4.7322E-01
LA-140	12.	1596.21	1.3433E-01
CE-141	282.	145.44	1.7506E-01
CE-144	224.	133.51	6.7954E-01
CEPR-144	224.	133.51	1.3578E+00
EU-152	266.	121.78	2.9482E-01
EU-154	27.	1274.51	4.4129E-01
EU-155	230.	86.54	3.4114E-01
HF-181	98.	482.18	1.4886E-01
TA-182	209.	67.75	4.8978E-01
HG-203	196.	279.20	1.3463E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL26980-BLNK

Acquisition date : 8-NOV-1998 22:18:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	80.	569.70	1.2775E-01
TL-208	197.	277.36	1.7335E+00
PB-210	402.	46.50	0.0000E+00
BI-212	63.	727.18	2.0771E+00
PB-212	250.	238.63	2.5256E-01
BI-214	164.	609.31	4.1548E-01
PB-214	217.	351.92	1.5156E+00
RA-224	259.	240.99	2.8720E+00
RA-226	265.	186.10	2.6318E+00
AC-228	99.	911.21	7.7627E-01
TH-228	249.	84.37	9.6193E+00
TH-229	209.	88.47	4.1629E-01
U-232	430.	57.78	3.3245E+02
PA-233	164.	312.17	2.8349E-01
UTH-233	216.	245.34	9.5149E+01
PA-234M	33.	1001.03	2.0761E+01
TH-234	215.	63.29	7.1477E+00
U-235	259.	185.71	1.5815E-01
NP-237	231.	86.48	9.0577E-01
U-237	399.	59.54	1.5038E+00
NP-238	46.	984.45	5.6413E-01
NP-239	233.	106.12	3.8130E-01
PU-239	229.	129.30	1.1930E+03
AM-241	399.	59.54	1.4372E+00
AM-243	255.	74.67	2.7200E-01

*****

* 222-S Laboratory Counting Room 9-NOV-1998 03:24:13.67 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: S98T2845-SAM
 Sample Size: 1.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

SL Hall

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3829.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:09.68 sec
 Dead Time: 0.1%

Verified by:

John Ralys 10 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 00:53:34.19
 Decayed to: 9-NOV-1998 00:53:34.19
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.74*	57598	1.59	1323.66	1315	18	0.9		CS-137	217.

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2845-SAM

Acquisition date : 9-NOV-1998 00:53:34

Total number of lines in spectrum	1
Number of unidentified lines	0
Number of lines tentatively identified by NID	1 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	2.170E+02	2.170E+02	0.019E+02	0.86	
Total Activity :			2.170E+02	2.170E+02			

Grand Total Activity :	2.170E+02	2.170E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2845-SAM

Acquisition date : 9-NOV-1998 00:53:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1940.	477.59	4.9029E+00
NA-24	25.	1368.55	1.6498E-01
K-40	365.	1460.75	5.3721E+00
AR-41	26.	1293.60	2.3653E-01
SC-46	65.	1120.55	2.0312E-01
CR-51	1604.	320.08	3.3611E+00
MN-54	58.	834.83	1.4985E-01
CO-56	45.	846.76	1.3459E-01
CO-58	67.	810.78	1.5622E-01
FE-59	34.	1099.25	2.6312E-01
CO-60	33.	1332.50	1.7126E-01
ZN-65	43.	1115.55	3.2814E-01
SE-75	1967.	264.66	5.5646E-01
KR-85	1038.	514.00	9.1060E+01
SR-85	1037.	514.01	4.1170E-01
Y-88	9.	1836.06	1.2416E-01
Y-91	37.	1204.67	5.5379E+01
NB-94	56.	871.09	1.5230E-01
NB-95	72.	765.78	1.5355E-01
ZR-95	55.	756.73	2.4615E-01
ZRNB-95	71.	724.18	6.6830E-01
RU-103	1216.	497.08	4.5746E-01
RURH-106	420.	621.93	6.0230E+00
AG-108m	74.	722.94	1.6314E-01
CD-109	2032.	88.03	8.7220E+00
AG-110M	1414.	657.76	5.9971E-01
SN-113	1712.	391.69	6.3604E-01
TE-123m	2145.	159.00	2.7840E-01
SB-124	411.	602.73	2.9444E-01
SB-125	2062.	427.89	1.6270E+00
TE-125m	2161.	109.27	8.8414E+01
I-129	1781.	39.60	0.0000E+00
I-131	1694.	364.48	4.7432E-01
XE-131m	2117.	163.93	1.2052E+01
BA-133	1645.	356.02	5.9852E-01
CS-134	386.	604.70	2.8690E-01
CS-136	57.	818.51	1.4670E-01
CS-138	13.	1435.86	5.3126E-01
CE-139	2127.	165.85	2.9728E-01
BA-140	732.	537.31	1.4203E+00
LA-140	10.	1596.21	1.2256E-01
CE-141	2075.	145.44	4.6521E-01
CE-144	2188.	133.51	2.0680E+00
CEPR-144	2188.	133.51	4.1320E+00
EU-152	2153.	121.78	8.2043E-01
EU-154	41.	1274.51	5.2859E-01
EU-155	2037.	86.54	9.9024E-01
HF-181	1613.	482.18	5.7999E-01
TA-182	1820.	67.75	1.4073E+00
HG-203	1870.	279.20	4.0485E-01
BI-207	483.	569.70	3.0316E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2845-SAM

Acquisition date : 9-NOV-1998 00:53:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1784.	277.36	5.0785E+00
PB-210	1879.	46.50	0.0000E+00
BI-212	80.	727.18	2.3235E+00
PB-212	2179.	238.63	7.2779E-01
BI-214	493.	609.31	7.0830E-01
PB-214	1782.	351.92	4.2418E+00
RA-224	2197.	240.99	8.1671E+00
RA-226	2542.	186.10	7.9603E+00
AC-228	99.	911.21	7.7705E-01
TH-228	2033.	84.37	2.6867E+01
TH-229	2065.	88.47	1.2742E+00
U-232	1993.	57.78	7.0559E+02
PA-233	1634.	312.17	8.6880E-01
UTH-233	2046.	245.34	2.8555E+02
PA-234M	48.	1001.03	2.4589E+01
TH-234	1789.	63.29	2.0120E+01
U-235	2510.	185.71	4.8078E-01
NP-237	2038.	86.48	2.6235E+00
U-237	1995.	59.54	3.3104E+00
NP-238	48.	984.45	5.7661E-01
NP-239	2012.	106.12	1.0939E+00
PU-239	2206.	129.30	3.6107E+03
AM-241	1995.	59.54	3.1640E+00
AM-243	1936.	74.67	7.3269E-01

*****

* 222-S Laboratory Counting Room 9-NOV-1998 06:08:11.96 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: S98T2845-DUP
 Sample Size: 1.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3830.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:09.83 sec
 Dead Time: 0.1%

Verified by:

John Reilly 12 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 03:37:27.36
 Decayed to: 9-NOV-1998 03:37:27.36
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.74*	58693	1.61	1323.66	1316	17	0.8		CS-137	221.

Summary of Nuclide Activity
 Sample ID : S98T2845-DUP

Page : 2
 Acquisition date : 9-NOV-1998 03:37:27

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma
			Uncorrected	Decay Corr		
			uCi/L	uCi/L	2-Sigma Error	%Error Flags
CS-137	30.00Y	1.00	2.211E+02	2.211E+02	0.019E+02	0.85
Total Activity :			2.211E+02	2.211E+02		

Grand Total Activity : 2.211E+02 2.211E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2845-DUP

Acquisition date : 9-NOV-1998 03:37:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2097.	477.59	5.0947E+00
NA-24	14.	1368.55	1.2712E-01
K-40	362.	1460.75	5.3520E+00
AR-41	27.	1293.60	2.3698E-01
SC-46	62.	1120.55	1.9913E-01
CR-51	1696.	320.08	3.4544E+00
MN-54	48.	834.83	1.3630E-01
CO-56	67.	846.76	1.6141E-01
CO-58	49.	810.78	1.3553E-01
FE-59	37.	1099.25	2.7219E-01
CO-60	32.	1332.50	1.6823E-01
ZN-65	43.	1115.55	3.2919E-01
SE-75	1904.	264.66	5.4750E-01
KR-85	1076.	514.00	9.2691E+01
SR-85	1075.	514.01	4.1914E-01
Y-88	8.	1836.06	1.1927E-01
Y-91	39.	1204.67	5.7090E+01
NB-94	45.	871.09	1.3833E-01
NB-95	59.	765.78	1.4046E-01
ZR-95	69.	756.73	2.7357E-01
ZRNB-95	85.	724.18	7.2609E-01
RU-103	1246.	497.08	4.6296E-01
RURH-106	430.	621.93	6.0871E+00
AG-108m	82.	722.94	1.7069E-01
CD-109	2107.	88.03	8.8790E+00
AG-110M	1415.	657.76	5.9978E-01
SN-113	1735.	391.69	6.4023E-01
TE-123m	2188.	159.00	2.8114E-01
SB-124	478.	602.73	3.1662E-01
SB-125	2048.	427.89	1.6216E+00
TE-125m	2098.	109.27	8.7132E+01
I-129	1894.	39.60	0.0000E+00
I-131	1714.	364.48	4.7712E-01
XE-131m	2141.	163.93	1.2119E+01
BA-133	1644.	356.02	5.9830E-01
CS-134	465.	604.70	3.1386E-01
CS-136	58.	818.51	1.4711E-01
CS-138	10.	1435.86	4.7474E-01
CE-139	2120.	165.85	2.9679E-01
BA-140	753.	537.31	1.4404E+00
LA-140	16.	1596.21	1.5407E-01
CE-141	2219.	145.44	4.8092E-01
CE-144	2153.	133.51	2.0515E+00
CEPR-144	2153.	133.51	4.0994E+00
EU-152	2237.	121.78	8.3614E-01
EU-154	24.	1274.51	4.1966E-01
EU-155	2121.	86.54	1.0103E+00
HF-181	1664.	482.18	5.8904E-01
TA-182	1885.	67.75	1.4319E+00
HG-203	1822.	279.20	3.9968E-01
BI-207	482.	569.70	3.0277E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2845-DUP

Acquisition date : 9-NOV-1998 03:37:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1839.	277.36	5.1550E+00
PB-210	1923.	46.50	0.0000E+00
BI-212	71.	727.18	2.1957E+00
PB-212	2177.	238.63	7.2741E-01
BI-214	505.	609.31	7.1639E-01
PB-214	1709.	351.92	4.1559E+00
RA-224	2193.	240.99	8.1584E+00
RA-226	2589.	186.10	8.0323E+00
AC-228	109.	911.21	8.1347E-01
TH-228	2052.	84.37	2.6993E+01
TH-229	2098.	88.47	1.2843E+00
U-232	2038.	57.78	7.1342E+02
PA-233	1630.	312.17	8.6776E-01
UTH-233	2108.	245.34	2.8981E+02
PA-234M	43.	1001.03	2.3411E+01
TH-234	1907.	63.29	2.0763E+01
U-235	2551.	185.71	4.8460E-01
NP-237	2126.	86.48	2.6785E+00
U-237	1989.	59.54	3.3054E+00
NP-238	41.	984.45	5.3747E-01
NP-239	2152.	106.12	1.1308E+00
PU-239	2221.	129.30	3.6224E+03
AM-241	1989.	59.54	3.1591E+00
AM-243	2005.	74.67	7.4556E-01

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 09:11:50.31 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26980
 Sample ID: S98T002918-SAM
 Sample Size: 5.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3831.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:10.79 sec
 Dead Time: 0.1%

Verified by:

10 NOV 98

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 9-NOV-1998 06:41:06.71
 Decayed to: 9-NOV-1998 06:41:06.71
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.74*	70169	1.60	1323.66	1315	18	0.8		CS-137	52.9
0	731.17	60	4.29	1462.49	1450	23	91.4			
0	1460.61*	36	2.16	2921.52	2914	171	29.8		K-40	0.418

Summary of Nuclide Activity
Sample ID : S98T002918-SAM

Page : 2
Acquisition date : 9-NOV-1998 06:41:06

Total number of lines in spectrum 3
Number of unidentified lines 1
Number of lines tentatively identified by NID 2 66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~4.177E-01~~	~~4.177E-01~~	~~5.421E-01~~	~~129.79~~	~~JFR~~
CS-137	30.00Y	1.00	5.287E+01	5.287E+01	0.041E+01	0.77	
Total Activity :			5.329E+01	5.329E+01			

11-10-98

Grand Total Activity : 5.329E+01 5.329E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002918-SAM

Acquisition date : 9-NOV-1998 06:41:06

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2389.	477.59	1.0866E+00
NA-24	19.	1368.55	2.9070E-02
AR-41	21.	1293.60	4.2385E-02
SC-46	74.	1120.55	4.3164E-02
CR-51	1949.	320.08	7.3992E-01
MN-54	63.	834.83	3.1017E-02
CO-56	57.	846.76	2.9953E-02
CO-58	49.	810.78	2.7083E-02
FE-59	29.	1099.25	4.8444E-02
CO-60	34.	1332.50	3.4597E-02
ZN-65	43.	1115.55	6.5699E-02
SE-75	2175.	264.66	1.1695E-01
KR-85	1240.	514.00	1.9880E+01
SR-85	1239.	514.01	8.9898E-02
Y-88	13.	1836.06	2.9292E-02
Y-91	43.	1204.67	1.1859E+01
NB-94	40.	871.09	2.6107E-02
NB-95	72.	765.78	3.0721E-02
ZR-95	47.	756.73	4.5926E-02
ZRNB-95	52.	724.18	1.1501E-01
RU-103	1470.	497.08	1.0044E-01
RURH-106	464.	621.93	1.2641E+00
AG-108m	58.	722.94	2.9094E-02
CD-109	1880.	88.03	1.6787E+00
AG-110M	1696.	657.76	1.3117E-01
SN-113	2115.	391.69	1.4117E-01
TE-123m	2190.	159.00	5.6243E-02
SB-124	493.	602.73	6.4310E-02
SB-125	2375.	427.89	3.4896E-01
TE-125m	1972.	109.27	1.6899E+01
I-129	1936.	39.60	0.0000E+00
I-131	1959.	364.48	1.0193E-01
XE-131m	2110.	163.93	2.4060E+00
BA-133	1921.	356.02	1.2921E-01
CS-134	483.	604.70	6.3970E-02
CS-136	53.	818.51	2.8208E-02
CS-138	17.	1435.86	1.1935E-01
CE-139	2195.	165.85	6.0390E-02
BA-140	836.	537.31	3.0326E-01
LA-140	16.	1596.21	3.0805E-02
CE-141	2177.	145.44	9.5278E-02
CE-144	2252.	133.51	4.1958E-01
CEPR-144	2252.	133.51	8.3839E-01
EU-152	2185.	121.78	1.6530E-01
EU-154	38.	1274.51	1.0234E-01
EU-155	1895.	86.54	1.9112E-01
HF-181	1843.	482.18	1.2388E-01
TA-182	1735.	67.75	2.7489E-01
HG-203	2048.	279.20	8.4680E-02
BI-207	583.	569.70	6.6461E-02
TL-208	2050.	277.36	1.0878E+00

4402

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002918-SAM

Acquisition date : 9-NOV-1998 06:41:06

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1955.	46.50	0.0000E+00
BI-212	79.	727.18	4.6156E-01
PB-212	2449.	238.63	1.5420E-01
BI-214	581.	609.31	1.5350E-01
PB-214	2053.	351.92	9.0983E-01
RA-224	2437.	240.99	1.7193E+00
RA-226	2843.	186.10	1.6824E+00
AC-228	86.	911.21	1.4514E-01
TH-228	1863.	84.37	5.1472E+00
TH-229	1860.	88.47	2.4204E-01
U-232	1961.	57.78	1.4002E+02
PA-233	1931.	312.17	1.8869E-01
UTH-233	2374.	245.34	6.1474E+01
PA-234M	45.	1001.03	4.7637E+00
TH-234	1815.	63.29	4.0525E+00
U-235	2830.	185.71	1.0202E-01
NP-237	1897.	86.48	5.0641E-01
U-237	1940.	59.54	6.5305E-01
NP-238	36.	984.45	1.0073E-01
NP-239	2035.	106.12	2.2003E-01
PU-239	2185.	129.30	7.1872E+02
AM-241	1940.	59.54	6.2416E-01
AM-243	1833.	74.67	1.4263E-01

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 11:58:35.16 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: S98T2918-DUP
 Sample Size: 5.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3832.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:09.31 sec
 Dead Time: 0.1%

Verified by:

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 09:27:58.66
 Decayed to: 9-NOV-1998 09:27:58.66
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.76*	58587	1.58	1323.69	1317	16	0.8		CS-137	44.1
0	911.41*	52	1.71	1822.94	1815	18	79.4			
0	1119.84*	29	2.41	2239.82	2234	11	91.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2918-DUP

Acquisition date : 9-NOV-1998 09:27:58

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	4.415E+01	4.415E+01	0.037E+01	0.84	
Total Activity :			4.415E+01	4.415E+01			

Grand Total Activity :	4.415E+01	4.415E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2918-DUP

Acquisition date : 9-NOV-1998 09:27:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1981.	477.59	9.9073E-01
NA-24	18.	1368.55	2.8257E-02
K-40	340.	1460.75	1.0374E+00
AR-41	28.	1293.60	4.8384E-02
SC-46	82.	1120.55	4.5291E-02
CR-51	1626.	320.08	6.7674E-01
MN-54	72.	834.83	3.3000E-02
CO-56	52.	846.76	2.8810E-02
CO-58	50.	810.78	2.7432E-02
FE-59	40.	1099.25	5.6411E-02
CO-60	20.	1332.50	2.7647E-02
ZN-65	36.	1115.55	6.0802E-02
SE-75	1897.	264.66	1.0930E-01
KR-85	1122.	514.00	1.8922E+01
SR-85	1121.	514.01	8.5571E-02
Y-88	6.	1836.06	2.0929E-02
Y-91	46.	1204.67	1.2222E+01
NB-94	36.	871.09	2.4789E-02
NB-95	53.	765.78	2.6593E-02
ZR-95	50.	756.73	4.6858E-02
ZRNB-95	61.	724.18	1.2450E-01
RU-103	1197.	497.08	9.0767E-02
RURH-106	401.	621.93	1.1776E+00
AG-108m	55.	722.94	2.8314E-02
CD-109	1636.	88.03	1.5678E+00
AG-110M	1437.	657.76	1.2086E-01
SN-113	1817.	391.69	1.3099E-01
TE-123m	1780.	159.00	5.0776E-02
SB-124	402.	602.73	5.8203E-02
SB-125	1976.	427.89	3.1868E-01
TE-125m	1700.	109.27	1.5707E+01
I-129	1707.	39.60	0.0000E+00
I-131	1567.	364.48	9.1289E-02
XE-131m	1941.	163.93	2.3094E+00
BA-133	1560.	356.02	1.1662E-01
CS-134	412.	604.70	5.9198E-02
CS-136	52.	818.51	2.7966E-02
CS-138	21.	1435.86	1.3090E-01
CE-139	1987.	165.85	5.7482E-02
BA-140	708.	537.31	2.7950E-01
LA-140	15.	1596.21	2.9860E-02
CE-141	1943.	145.44	9.0069E-02
CE-144	1891.	133.51	3.8485E-01
CEPR-144	1891.	133.51	7.6908E-01
EU-152	1830.	121.78	1.5146E-01
EU-154	33.	1274.51	9.6554E-02
EU-155	1708.	86.54	1.8156E-01
HF-181	1536.	482.18	1.1323E-01
TA-182	1448.	67.75	2.5148E-01
HG-203	1688.	279.20	7.6969E-02
BI-207	496.	569.70	6.1446E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2918-DUP

Acquisition date : 9-NOV-1998 09:27:58

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1764.	277.36	1.0101E+00
PB-210	1692.	46.50	0.0000E+00
BI-212	79.	727.18	4.6167E-01
PB-212	2090.	238.63	1.4259E-01
BI-214	480.	609.31	1.3986E-01
PB-214	1665.	351.92	8.2037E-01
RA-224	2028.	240.99	1.5699E+00
RA-226	2389.	186.10	1.5438E+00
AC-228	112.	911.21	1.6460E-01
TH-228	1659.	84.37	4.8604E+00
TH-229	1617.	88.47	2.2585E-01
U-232	1682.	57.78	1.2980E+02
PA-233	1656.	312.17	1.7489E-01
UTH-233	2030.	245.34	5.6892E+01
PA-234M	53.	1001.03	5.1441E+00
TH-234	1513.	63.29	3.7052E+00
U-235	2376.	185.71	9.3574E-02
NP-237	1707.	86.48	4.8067E-01
U-237	1691.	59.54	6.1017E-01
NP-238	39.	984.45	1.0432E-01
NP-239	1672.	106.12	1.9967E-01
PU-239	1882.	129.30	6.6771E+02
AM-241	1691.	59.54	5.8319E-01
AM-243	1708.	74.67	1.3775E-01

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 15:09:39.52 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 26980
 Sample ID: S98T2919-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3834.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.08 sec
 Dead Time: 0.1%

Verified by:

John Relyea 10 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 9-NOV-1998 12:39:02.29
 Decayed to: 9-NOV-1998 12:39:02.29
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	474.60	140	3.31	949.48	944	9100.6				
0	661.75*	40576	1.58	1323.68	1315	18	1.0		CS-137	15.3

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2919-SAM

Acquisition date : 9-NOV-1998 12:39:02

Total number of lines in spectrum	2	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	1	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.529E+01	1.529E+01	0.016E+01	1.02	
Total Activity :			1.529E+01	1.529E+01			

Grand Total Activity :	1.529E+01	1.529E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2919-SAM

Acquisition date : 9-NOV-1998 12:39:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1384.	477.59	4.1506E-01
NA-24	20.	1368.55	1.4935E-02
K-40	364.	1460.75	5.3660E-01
AR-41	27.	1293.60	2.4016E-02
SC-46	61.	1120.55	1.9780E-02
CR-51	1171.	320.08	2.8784E-01
MN-54	47.	834.83	1.3593E-02
CO-56	39.	846.76	1.2613E-02
CO-58	56.	810.78	1.4363E-02
FE-59	37.	1099.25	2.7342E-02
CO-60	29.	1332.50	1.6072E-02
ZN-65	34.	1115.55	2.9591E-02
SE-75	1316.	264.66	4.5648E-02
KR-85	826.	514.00	8.1428E+00
SR-85	826.	514.01	3.6822E-02
Y-88	7.	1836.06	1.1288E-02
Y-91	50.	1204.67	6.3422E+00
NB-94	53.	871.09	1.4882E-02
NB-95	71.	765.78	1.5337E-02
ZR-95	60.	756.73	2.5591E-02
ZRNB-95	64.	724.18	6.3661E-02
RU-103	864.	497.08	3.8671E-02
RURH-106	310.	621.93	5.1921E-01
AG-108m	59.	722.94	1.4664E-02
CD-109	1204.	88.03	6.7400E-01
AG-110M	1003.	657.76	5.0644E-02
SN-113	1302.	391.69	5.5576E-02
TE-123m	1438.	159.00	2.2858E-02
SB-124	315.	602.73	2.5862E-02
SB-125	1390.	427.89	1.3395E-01
TE-125m	1202.	109.27	6.6221E+00
I-129	1317.	39.60	0.0000E+00
I-131	1210.	364.48	4.0197E-02
XE-131m	1363.	163.93	9.7009E-01
BA-133	1119.	356.02	4.9523E-02
CS-134	329.	604.70	2.6544E-02
CS-136	41.	818.51	1.2588E-02
CS-138	22.	1435.86	6.6927E-02
CE-139	1383.	165.85	2.4043E-02
BA-140	504.	537.31	1.1838E-01
LA-140	10.	1596.21	1.2713E-02
CE-141	1361.	145.44	3.7794E-02
CE-144	1387.	133.51	1.6517E-01
CEPR-144	1387.	133.51	3.3004E-01
EU-152	1335.	121.78	6.4826E-02
EU-154	40.	1274.51	5.2214E-02
EU-155	1178.	86.54	7.5589E-02
HF-181	1124.	482.18	4.8553E-02
TA-182	1054.	67.75	1.0752E-01
HG-203	1340.	279.20	3.4356E-02
BI-207	336.	569.70	2.5414E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2919-SAM

Acquisition date : 9-NOV-1998 12:39:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1314.	277.36	4.3687E-01
PB-210	1304.	46.50	0.0000E+00
BI-212	83.	727.18	2.3611E-01
PB-212	1536.	238.63	6.1245E-02
BI-214	387.	609.31	6.2912E-02
PB-214	1253.	351.92	3.5651E-01
RA-224	1462.	240.99	6.6804E-01
RA-226	1725.	186.10	6.5730E-01
AC-228	83.	911.21	7.1671E-02
TH-228	1178.	84.37	2.0536E+00
TH-229	1212.	88.47	9.8003E-02
U-232	1240.	57.78	5.5852E+01
PA-233	1190.	312.17	7.4310E-02
UTH-233	1502.	245.34	2.4520E+01
PA-234M	58.	1001.03	2.6779E+00
TH-234	1132.	63.29	1.6057E+00
U-235	1687.	185.71	3.9509E-02
NP-237	1177.	86.48	2.0013E-01
U-237	1277.	59.54	2.6568E-01
NP-238	30.	984.45	4.6288E-02
NP-239	1214.	106.12	8.5279E-02
PU-239	1339.	129.30	2.8230E+02
AM-241	1277.	59.54	2.5393E-01
AM-243	1156.	74.67	5.6841E-02

*****
 * 222-S Laboratory Counting Room 9-NOV-1998 18:10:26.51 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 26980
 Sample ID: S98T002919-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Shelley

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3835.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:06.95 sec
 Dead Time: 0.1%

Verified by:

John Relyea 10 Nov 98

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 9-NOV-1998 15:39:29.06
 Decayed to: 9-NOV-1998 15:39:29.06
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.75*	39297	1.57	1323.68	1316	17	1.0		CS-137	14.8
0	910.66*	44	4.16	1821.44	1813	19	87.3			

Summary of Nuclide Activity

Sample ID : S98T002919-DUP

Page : 2

Acquisition date : 9-NOV-1998 15:39:29

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.481E+01	1.481E+01	0.015E+01	1.04	
Total Activity :			1.481E+01	1.481E+01			

Grand Total Activity :	1.481E+01	1.481E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002919-DUP

Acquisition date : 9-NOV-1998 15:39:29

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1338.	477.59	4.0829E-01
NA-24	13.	1368.55	1.2344E-02
K-40	362.	1460.75	5.3501E-01
AR-41	26.	1293.60	2.3336E-02
SC-46	71.	1120.55	2.1211E-02
CR-51	1109.	320.08	2.8025E-01
MN-54	49.	834.83	1.3815E-02
CO-56	54.	846.76	1.4670E-02
CO-58	45.	810.78	1.3036E-02
FE-59	47.	1099.25	3.0493E-02
CO-60	30.	1332.50	1.6333E-02
ZN-65	36.	1115.55	3.0230E-02
SE-75	1304.	264.66	4.5438E-02
KR-85	779.	514.00	7.9111E+00
SR-85	778.	514.01	3.5772E-02
Y-88	6.	1836.06	1.0602E-02
Y-91	42.	1204.67	5.8524E+00
NB-94	44.	871.09	1.3660E-02
NB-95	57.	765.78	1.3791E-02
ZR-95	64.	756.73	2.6369E-02
ZRNB-95	70.	724.18	6.6350E-02
RU-103	840.	497.08	3.8153E-02
RURH-106	314.	621.93	5.2303E-01
AG-108m	70.	722.94	1.5901E-02
CD-109	1154.	88.03	6.5994E-01
AG-110M	937.	657.76	4.8987E-02
SN-113	1176.	391.69	5.2870E-02
TE-123m	1385.	159.00	2.2434E-02
SB-124	315.	602.73	2.5872E-02
SB-125	1370.	427.89	1.3298E-01
TE-125m	1152.	109.27	6.4841E+00
I-129	1256.	39.60	0.0000E+00
I-131	1172.	364.48	3.9573E-02
XE-131m	1316.	163.93	9.5327E-01
BA-133	1122.	356.02	4.9584E-02
CS-134	331.	604.70	2.6620E-02
CS-136	44.	818.51	1.3061E-02
CS-138	15.	1435.86	5.6491E-02
CE-139	1285.	165.85	2.3188E-02
BA-140	525.	537.31	1.2072E-01
LA-140	16.	1596.21	1.5224E-02
CE-141	1370.	145.44	3.7917E-02
CE-144	1294.	133.51	1.5959E-01
CEPR-144	1294.	133.51	3.1893E-01
EU-152	1319.	121.78	6.4452E-02
EU-154	27.	1274.51	4.4182E-02
EU-155	1177.	86.54	7.5584E-02
HF-181	1115.	482.18	4.8359E-02
TA-182	1106.	67.75	1.1012E-01
HG-203	1228.	279.20	3.2900E-02
BI-207	348.	569.70	2.5865E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002919-DUP

Acquisition date : 9-NOV-1998 15:39:29

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1271.	277.36	4.2981E-01
PB-210	1339.	46.50	0.0000E+00
BI-212	67.	727.18	2.1367E-01
PB-212	1540.	238.63	6.1323E-02
BI-214	383.	609.31	6.2624E-02
PB-214	1171.	351.92	3.4482E-01
RA-224	1512.	240.99	6.7911E-01
RA-226	1731.	186.10	6.5838E-01
AC-228	87.	911.21	7.2979E-02
TH-228	1172.	84.37	2.0481E+00
TH-229	1168.	88.47	9.6228E-02
U-232	1271.	57.78	5.6531E+01
PA-233	1128.	312.17	7.2381E-02
UTH-233	1419.	245.34	2.3846E+01
PA-234M	38.	1001.03	2.2222E+00
TH-234	1026.	63.29	1.5304E+00
U-235	1722.	185.71	3.9913E-02
NP-237	1178.	86.48	2.0025E-01
U-237	1232.	59.54	2.6102E-01
NP-238	39.	984.45	5.2557E-02
NP-239	1210.	106.12	8.5138E-02
PU-239	1270.	129.30	2.7496E+02
AM-241	1232.	59.54	2.4948E-01
AM-243	1171.	74.67	5.7198E-02

LABCORE Completed Worklist Report for Worklist# 27141

Analyst: scl

Instrument: GEA02

Book#: 28B57

Method: LA-548-121

Rev/Mod F-φ

Worklist Comment: BY105, @GEA-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	@GEA-01 CO60-02	SOLID	4.517e-03	4.71e-03	104.273	% Recovery
1 STD	0	0	@GEA-01 CO60-02H	SOLID	1	3.16	3.160	% Ct Error
1 STD	0	0	@GEA-01 CS13702	SOLID	7.449e-03	7.62e-03	102.296	% Recovery
1 STD	0	0	@GEA-01 CS13702H	SOLID	1	3.09	3.090	% Ct Error
2 BLNK-PREP	0	0	@GEA-01 CO60-02	SOLID	1	<4.14e-2		uCi/g
2 BLNK-PREP	0	0	@GEA-01 CS13702	SOLID	1	1.87e-01	0.187	uCi/g
3 SAMPLE	S98T002712	0 F	@GEA-01 CO60-02	SOLID	N/A	4.082e-02	4.08e-002	uCi/g
3 SAMPLE	S98T002712	0 F	@GEA-01 CO60-02H	SOLID	N/A	n/a		% Ct. Error
3 SAMPLE	S98T002712	0 F	@GEA-01 CS13702	SOLID	N/A	1.119e+02		uCi/g
3 SAMPLE	S98T002712	0 F	@GEA-01 CS13702H	SOLID	N/A	0.610		% Ct. Error
4 DUP	S98T002712	0 F	@GEA-01 CO60-02	SOLID	<4.08e-2	<4.45e-2		RPD
4 DUP	S98T002712	0 F	@GEA-01 CO60-02H	SOLID	1	n/a		% Ct Error
4 DUP	S98T002712	0 F	@GEA-01 CS13702	SOLID	1.12e+02	9.28e+01	18.750	RPD
4 DUP	S98T002712	0 F	@GEA-01 CS13702H	SOLID	1.00	0.670	0.670	% Ct Error
5 SAMPLE	S98T002739	0 F	@GEA-01 CO60-02	SOLID	N/A	1.014e-02	1.01e-002	uCi/g
5 SAMPLE	S98T002739	0 F	@GEA-01 CO60-02H	SOLID	N/A	n/a		% Ct. Error
5 SAMPLE	S98T002739	0 F	@GEA-01 CS13702	SOLID	N/A	2.763e+01		uCi/g
5 SAMPLE	S98T002739	0 F	@GEA-01 CS13702H	SOLID	N/A	0.610		% Ct. Error
6 DUP	S98T002739	0 F	@GEA-01 CO60-02	SOLID	<1.01e-2	<1.03e-2		RPD
6 DUP	S98T002739	0 F	@GEA-01 CO60-02H	SOLID	1	n/a		% Ct Error
6 DUP	S98T002739	0 F	@GEA-01 CS13702	SOLID	2.76e+01	2.88e+01	4.295	RPD
6 DUP	S98T002739	0 F	@GEA-01 CS13702H	SOLID	1.00	0.590	0.590	% Ct Error
7 SAMPLE	S98T002844	0 F	@GEA-01 CO60-02	SOLID	N/A	2.289e-02	2.29e-002	uCi/g
7 SAMPLE	S98T002844	0 F	@GEA-01 CO60-02H	SOLID	N/A	n/a		% Ct. Error
7 SAMPLE	S98T002844	0 F	@GEA-01 CS13702	SOLID	N/A	6.548e+01		uCi/g
7 SAMPLE	S98T002844	0 F	@GEA-01 CS13702H	SOLID	N/A	0.560		% Ct. Error
8 DUP	S98T002844	0 F	@GEA-01 CO60-02	SOLID	<2.29e-2	<1.87e-2		RPD
8 DUP	S98T002844	0 F	@GEA-01 CO60-02H	SOLID	1	n/a		% Ct Error
8 DUP	S98T002844	0 F	@GEA-01 CS13702	SOLID	6.55e+01	6.32e+01	5.494	RPD
8 DUP	S98T002844	0 F	@GEA-01 CS13702H	SOLID	1.00	0.540	0.540	% Ct Error

Comments Section:

Comments for sample# S98T002712 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002739 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002844 and test @GEA-01 .

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27141

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit

DL=0 => n/a.

Final page for worklist# 27141

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature

11-23-98
Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

11/11/98 18:32
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 27141

Analyst: S.L Instrument: GEA00 Book# 21857Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105, @GEA-01, tdm

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T002712	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E								
4	DUP	S98T002712	0	F	@GEA-01	SOLID		
5	SAMPLE	S98T002739	0	F	@GEA-01	SOLID	98000490	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E								
6	DUP	S98T002739	0	F	@GEA-01	SOLID		
7	SAMPLE	S98T002844	0	F	@GEA-01	SOLID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E								
8	DUP	S98T002844	0	F	@GEA-01	SOLID		

Final page for worklist # 27141

Sue Lai 11-20-98
 Analyst Signature Date

Shirley Hale 11-23-98
 Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 11:44:01.07 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27141
 Sample ID: WL27141-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2862.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:22.93 sec
 Dead Time: 0.8%

Verified by:

23/Nov98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 10:53:08.86
 Decayed to: 20-NOV-1998 10:53:08.86
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.49*	5240	1.45	1322.98	1315	16	3.1		CS-137	7.62
0	1172.92	2401	1.80	2345.88	2337	15	4.4		CO-60	4.76
0	1332.17	2118	1.86	2664.42	2654	20	4.5		CO-60	4.67

Summary of Nuclide Activity

Page : 2

Sample ID : WL27141-STD

Acquisition date : 20-NOV-1998 10:53:08

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected uCi/L	Decay Corr uCi/L			
CO-60	5.27Y	1.00	4.714E+00	4.714E+00	0.149E+00	3.16	
CS-137	30.00Y	1.00	7.621E+00	7.621E+00	0.236E+00	3.09	
Total Activity :			1.234E+01	1.234E+01			

Grand Total Activity :	1.234E+01	1.234E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27141-STD

Acquisition date : 20-NOV-1998 10:53:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	248.	477.59	6.7825E-01
NA-24	6.	1368.55	3.2381E-02
K-40	113.	1460.75	1.1662E+00
AR-41	22.	1293.60	6.1967E-02
SC-46	90.	1120.55	8.9460E-02
CR-51	251.	320.08	4.8289E-01
MN-54	100.	834.83	7.4202E-02
CO-56	138.	846.76	8.7411E-02
CO-58	146.	810.78	8.7158E-02
FE-59	125.	1099.25	1.8219E-01
ZN-65	115.	1115.55	1.9681E-01
SE-75	281.	264.66	7.4522E-02
KR-85	188.	514.00	1.5164E+01
SR-85	188.	514.01	6.8554E-02
Y-88	7.	1836.06	4.5464E-02
Y-91	35.	1204.67	2.0506E+01
NB-94	143.	871.09	9.1072E-02
NB-95	116.	765.78	7.4160E-02
ZR-95	105.	756.73	1.2847E-01
ZRNB-95	96.	724.18	2.9640E-01
RU-103	198.	497.08	7.1870E-02
RURH-106	131.	621.93	1.3225E+00
AG-108m	96.	722.94	7.1141E-02
CD-109	279.	88.03	1.3476E+00
AG-110M	241.	657.76	9.7674E-02
SN-113	266.	391.69	9.4228E-02
TE-123m	307.	159.00	4.1163E-02
SB-124	135.	602.73	6.6381E-02
SB-125	291.	427.89	2.3259E-01
TE-125m	292.	109.27	1.3437E+01
I-129	275.	39.60	1.5385E+01
I-131	265.	364.48	6.9338E-02
XE-131m	336.	163.93	1.8533E+00
BA-133	246.	356.02	8.5559E-02
CS-134	123.	604.70	6.3772E-02
CS-136	133.	818.51	8.3763E-02
CS-138	9.	1435.86	8.3232E-02
CE-139	329.	165.85	4.5080E-02
BA-140	159.	537.31	2.5974E-01
LA-140	9.	1596.21	4.5555E-02
CE-141	330.	145.44	7.3861E-02
CE-144	329.	133.51	3.2469E-01
CEPR-144	329.	133.51	6.4861E-01
EU-152	314.	121.78	1.2832E-01
EU-154	18.	1274.51	1.3958E-01
EU-155	299.	86.54	1.5795E-01
HF-181	213.	482.18	8.1792E-02
TA-182	276.	67.75	2.1542E-01
HG-203	324.	279.20	5.9858E-02
BI-207	139.	569.70	6.3934E-02
TL-208	309.	277.36	7.5120E-01

4421

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27141-STD

Acquisition date : 20-NOV-1998 10:53:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	283.	46.50	1.0749E+01
BI-212	110.	727.18	1.0404E+00
PB-212	395.	238.63	1.0877E-01
BI-214	159.	609.31	1.5811E-01
PB-214	315.	351.92	2.9567E-01
RA-224	377.	240.99	1.1801E+00
RA-226	357.	186.10	1.1125E+00
AC-228	165.	911.21	3.7983E-01
TH-228	291.	84.37	4.2254E+00
TH-229	275.	88.47	1.9395E-01
U-232	270.	57.78	7.6271E+01
PA-233	278.	312.17	1.2939E-01
UTH-233	361.	245.34	4.2144E+01
PA-234M	126.	1001.03	1.4747E+01
TH-234	262.	63.29	2.8320E+00
U-235	362.	185.71	6.8117E-02
NP-237	299.	86.48	4.1806E-01
U-237	307.	59.54	4.1992E-01
NP-238	131.	984.45	3.4762E-01
NP-239	318.	106.12	1.7786E-01
PU-239	327.	129.30	5.6533E+02
AM-241	307.	59.54	4.0279E-01
AM-243	289.	74.67	1.1603E-01

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 14:42:44.25
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27141
 Sample ID: WL27141-BLK
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

Im Daula

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2864.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:37.60 sec
 Dead Time: 0.4%

Verified by:

John Riley 23 Nov 98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 12:11:35.95
 Decayed to: 20-NOV-1998 12:11:35.95
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	34.53	162	2.45	69.13	64	10	59.3			
0	295.26*	42	1.95	590.55	587	8111.3				
0	661.64*	190	1.29	1323.27	1318	11	28.9		CS-137	0.368

Summary of Nuclide Activity

Page : 2

Sample ID : WL27141-BLK

Acquisition date : 20-NOV-1998 12:11:35

Total number of lines in spectrum	3	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	2	66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	3.676E-01	3.676E-01	1.062E-01	28.89	
Total Activity :			3.676E-01	3.676E-01			

Grand Total Activity :	3.676E-01	3.676E-01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
Sample ID : WL27141-BLK

HNF-1663 REV.0

Page : 3
Acquisition date : 20-NOV-1998 12:11:35

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	101.	477.59	5.8844E-01
NA-24	19.	1368.55	7.3450E-02
K-40	309.	1460.75	2.5167E+00
AR-41	26.	1293.60	1.1881E-01
SC-46	68.	1120.55	1.0425E-01
CR-51	126.	320.08	4.6272E-01
MN-54	46.	834.83	6.9035E-02
CO-56	46.	846.76	6.9542E-02
CO-58	29.	810.78	5.5116E-02
FE-59	27.	1099.25	1.1850E-01
CO-60	29.	1332.50	8.1390E-02
ZN-65	29.	1115.55	1.3803E-01
SE-75	157.	264.66	7.4983E-02
KR-85	118.	514.00	1.6155E+01
SR-85	117.	514.01	7.2961E-02
Y-88	10.	1836.06	6.9734E-02
Y-91	39.	1204.67	2.8585E+01
NB-94	28.	871.09	5.6653E-02
NB-95	44.	765.78	6.3063E-02
ZR-95	60.	756.73	1.3115E-01
ZRNB-95	47.	724.18	2.8315E-01
RU-103	70.	497.08	5.8500E-02
RURH-106	58.	621.93	1.2004E+00
AG-108m	50.	722.94	6.9574E-02
CD-109	186.	88.03	1.4792E+00
AG-110M	60.	657.76	6.7325E-02
SN-113	112.	391.69	8.3195E-02
TE-123m	222.	159.00	4.6884E-02
SB-124	61.	602.73	6.1007E-02
SB-125	132.	427.89	2.1209E-01
TE-125m	188.	109.27	1.4472E+01
I-129	163.	39.60	1.5971E+01
I-131	135.	364.48	6.7084E-02
XE-131m	235.	163.93	2.0842E+00
BA-133	108.	356.02	7.6883E-02
CS-134	63.	604.70	6.2076E-02
CS-136	48.	818.51	6.9704E-02
CS-138	21.	1435.86	3.3280E-01
CE-139	218.	165.85	4.9266E-02
BA-140	59.	537.31	2.1641E-01
LA-140	6.	1596.21	5.1079E-02
CE-141	195.	145.44	7.6551E-02
CE-144	218.	133.51	3.5501E-01
CEPR-144	218.	133.51	7.0930E-01
EU-152	179.	121.78	1.3067E-01
EU-154	32.	1274.51	2.3712E-01
EU-155	188.	86.54	1.6838E-01
HF-181	94.	482.18	7.3747E-02
TA-182	169.	67.75	2.2671E-01
HG-203	167.	279.20	5.8083E-02
BI-207	83.	569.70	6.6588E-02

4425

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27141-BLK

Acquisition date : 20-NOV-1998 12:11:35

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	181.	277.36	7.7418E-01
PB-210	169.	46.50	1.1210E+01
BI-212	59.	727.18	1.0387E+00
PB-212	240.	238.63	1.1394E-01
BI-214	148.	609.31	2.0360E-01
PB-214	211.	351.92	7.2225E-01
RA-224	230.	240.99	1.2467E+00
RA-226	261.	186.10	1.2758E+00
AC-228	75.	911.21	3.4719E-01
TH-228	216.	84.37	4.8891E+00
TH-229	186.	88.47	2.1391E-01
U-232	167.	57.78	8.0637E+01
PA-233	123.	312.17	1.1684E-01
UTH-233	191.	245.34	4.1282E+01
PA-234M	41.	1001.03	1.1651E+01
TH-234	162.	63.29	2.9971E+00
U-235	264.	185.71	7.7923E-02
NP-237	190.	86.48	4.4789E-01
U-237	173.	59.54	4.2604E-01
NP-238	32.	984.45	2.4327E-01
NP-239	176.	106.12	1.8043E-01
PU-239	216.	129.30	6.1706E+02
AM-241	173.	59.54	4.0722E-01
AM-243	223.	74.67	1.3641E-01

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 17:24:25.08 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27141
 Sample ID: S98T2712-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2865.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:28.41 sec
 Dead Time: 1.0%

Verified by:

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 20-NOV-1998 14:52:26.99
 Decayed to: 20-NOV-1998 14:52:26.99
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	661.45*	113352	1.47	1322.89	1315	16 0.6		CS-137	220.
0	1460.44*	24	2.03	2921.01	2914	15175.2		K-40	0.703
0	1592.29	12	1.02	3184.75	3181	9107.0			
0	1763.93*	17	1.95	3528.08	3520	15 98.6			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2712-SAM

Acquisition date : 20-NOV-1998 14:52:26

Total number of lines in spectrum 4
 Number of unidentified lines 1
 Number of lines tentatively identified by NID 3 75.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~7.030E-01~~	~~7.030E-01~~	~~12.31E-01~~	~~175.15~~	~~1123198~~
CS-137	30.00Y	1.00	2.198E+02	2.198E+02	0.013E+02	0.61	1123198
Total Activity :			2.205E+02	2.205E+02			

Grand Total Activity : 2.205E+02 2.205E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2712-SAM

Acquisition date : 20-NOV-1998 14:52:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	3542.	477.59	3.3307E+00
NA-24	18.	1368.55	7.1762E-02
AR-41	22.	1293.60	1.0881E-01
SC-46	72.	1120.55	1.0704E-01
CR-51	2945.	320.08	2.1502E+00
MN-54	54.	834.83	7.4208E-02
CO-56	37.	846.76	6.2975E-02
CO-58	36.	810.78	6.0273E-02
FE-59	44.	1099.25	1.4841E-01
CO-60	28.	1332.50	8.0282E-02
ZN-65	30.	1115.55	1.4054E-01
SE-75	3251.	264.66	3.2983E-01
KR-85	1803.	514.00	6.0848E+01
SR-85	1802.	514.01	2.7519E-01
Y-88	11.	1836.06	7.2464E-02
Y-91	35.	1204.67	2.7183E+01
NB-94	37.	871.09	6.4121E-02
NB-95	63.	765.78	7.4449E-02
ZR-95	51.	756.73	1.2247E-01
ZRNB-95	71.	724.18	3.4247E-01
RU-103	2091.	497.08	3.0299E-01
RURH-106	761.	621.93	4.1355E+00
AG-108m	63.	722.94	7.7747E-02
CD-109	3044.	88.03	5.7936E+00
AG-110M	3174.	657.76	4.6010E-01
SN-113	3041.	391.69	4.1455E-01
TE-123m	3392.	159.00	1.7826E-01
SB-124	822.	602.73	2.1194E-01
SB-125	3659.	427.89	1.0744E+00
TE-125m	3198.	109.27	5.7903E+01
I-129	2952.	39.60	6.5674E+01
I-131	2926.	364.48	3.0088E-01
XE-131m	3328.	163.93	7.6309E+00
BA-133	3061.	356.02	3.9217E-01
CS-134	817.	604.70	2.1226E-01
CS-136	47.	818.51	6.8857E-02
CS-138	26.	1435.86	3.7028E-01
CE-139	3465.	165.85	1.9095E-01
BA-140	1287.	537.31	9.5774E-01
LA-140	13.	1596.21	7.1324E-02
CE-141	3482.	145.44	3.1335E-01
CE-144	3558.	133.51	1.3925E+00
CEPR-144	3558.	133.51	2.7825E+00
EU-152	3400.	121.78	5.5053E-01
EU-154	39.	1274.51	2.5961E-01
EU-155	3057.	86.54	6.5811E-01
HF-181	2791.	482.18	3.8407E-01
TA-182	2830.	67.75	8.9871E-01
HG-203	3202.	279.20	2.4585E-01
BI-207	915.	569.70	2.1230E-01
TL-208	3162.	277.36	3.1345E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2712-SAM

Acquisition date : 20-NOV-1998 14:52:26

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	2730.	46.50	4.3647E+01
BI-212	70.	727.18	1.1242E+00
PB-212	3780.	238.63	4.4006E-01
BI-214	880.	609.31	4.8278E-01
PB-214	3129.	351.92	2.7171E+00
RA-224	3708.	240.99	4.8697E+00
RA-226	4370.	186.10	5.0810E+00
AC-228	83.	911.21	3.6572E-01
TH-228	3063.	84.37	1.7878E+01
TH-229	3068.	88.47	8.4295E-01
U-232	2740.	57.78	3.1641E+02
PA-233	2945.	312.17	5.4901E-01
UTH-233	3638.	245.34	1.7468E+02
PA-234M	28.	1001.03	9.6872E+00
TH-234	2833.	63.29	1.2113E+01
U-235	4288.	185.71	3.0612E-01
NP-237	3060.	86.48	1.7439E+00
U-237	2725.	59.54	1.6383E+00
NP-238	44.	984.45	2.8187E-01
NP-239	3077.	106.12	7.2954E-01
PU-239	3469.	129.30	2.4018E+03
AM-241	2725.	59.54	1.5659E+00
AM-243	2999.	74.67	4.8672E-01

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 20:52:11.02
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27141
 Sample ID: S98T002712-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2866.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:21.51 sec
 Dead Time: 0.9%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 18:20:19.92
 Decayed to: 20-NOV-1998 18:20:19.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.45*	93935	1.47	1322.89	1315	16	0.7		CS-137	182.
0	1460.40*	27	1.62	2920.92	2912	16152.9			K-40	0.813
0	1763.67*	13	2.06	3527.56	3519	17111.7				

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002712-DUP

Acquisition date : 20-NOV-1998 18:20:19

Total number of lines in spectrum 3
 Number of unidentified lines 0
 Number of lines tentatively identified by NID 3 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~8.135E-01~~	~~8.135E-01~~	~~12.44E-01~~	~~152.94~~	~~11/23/98~~
CS-137	30.00Y	1.00	1.822E+02	1.822E+02	0.012E+02	0.67	
Total Activity :			1.830E+02	1.830E+02			

Grand Total Activity : 1.830E+02 1.830E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002712-DUP

Acquisition date : 20-NOV-1998 18:20:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2873.	477.59	3.0028E+00
NA-24	15.	1368.55	6.6752E-02
AR-41	24.	1293.60	1.1490E-01
SC-46	61.	1120.55	9.9397E-02
CR-51	2362.	320.08	1.9281E+00
MN-54	48.	834.83	7.0262E-02
CO-56	42.	846.76	6.7100E-02
CO-58	42.	810.78	6.4941E-02
FE-59	33.	1099.25	1.3033E-01
CO-60	34.	1332.50	8.7368E-02
ZN-65	34.	1115.55	1.4912E-01
SE-75	2774.	264.66	3.0490E-01
KR-85	1556.	514.00	5.6581E+01
SR-85	1556.	514.01	2.5592E-01
Y-88	11.	1836.06	7.2783E-02
Y-91	29.	1204.67	2.5137E+01
NB-94	44.	871.09	6.9450E-02
NB-95	48.	765.78	6.5588E-02
ZR-95	50.	756.73	1.2137E-01
ZRNB-95	58.	724.18	3.1169E-01
RU-103	1739.	497.08	2.7667E-01
RURH-106	649.	621.93	3.8260E+00
AG-108m	59.	722.94	7.5520E-02
CD-109	2618.	88.03	5.3766E+00
AG-110M	2614.	657.76	4.1794E-01
SN-113	2639.	391.69	3.8646E-01
TE-123m	2909.	159.00	1.6519E-01
SB-124	723.	602.73	1.9905E-01
SB-125	3081.	427.89	9.8668E-01
TE-125m	2784.	109.27	5.4071E+01
I-129	2312.	39.60	5.8192E+01
I-131	2522.	364.48	2.7960E-01
XE-131m	2826.	163.93	7.0378E+00
BA-133	2445.	356.02	3.5093E-01
CS-134	716.	604.70	1.9904E-01
CS-136	46.	818.51	6.8327E-02
CS-138	15.	1435.86	2.9204E-01
CE-139	2879.	165.85	1.7424E-01
BA-140	1113.	537.31	8.9173E-01
LA-140	11.	1596.21	6.8101E-02
CE-141	2972.	145.44	2.8973E-01
CE-144	2824.	133.51	1.2421E+00
CEPR-144	2824.	133.51	2.4821E+00
EU-152	2923.	121.78	5.1083E-01
EU-154	32.	1274.51	2.3698E-01
EU-155	2587.	86.54	6.0600E-01
HF-181	2285.	482.18	3.4792E-01
TA-182	2463.	67.75	8.3907E-01
HG-203	2680.	279.20	2.2515E-01
BI-207	731.	569.70	1.9028E-01
TL-208	2715.	277.36	2.9065E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002712-DUP

Acquisition date : 20-NOV-1998 18:20:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	2370.	46.50	4.0696E+01
BI-212	65.	727.18	1.0804E+00
PB-212	3147.	238.63	4.0186E-01
BI-214	756.	609.31	4.4830E-01
PB-214	2555.	351.92	2.4561E+00
RA-224	3059.	240.99	4.4273E+00
RA-226	3682.	186.10	4.6674E+00
AC-228	71.	911.21	3.3960E-01
TH-228	2598.	84.37	1.6479E+01
TH-229	2589.	88.47	7.7506E-01
U-232	2236.	57.78	2.8622E+02
PA-233	2473.	312.17	5.0357E-01
UTH-233	3089.	245.34	1.6110E+02
PA-234M	55.	1001.03	1.3264E+01
TH-234	2322.	63.29	1.0979E+01
U-235	3644.	185.71	2.8237E-01
NP-237	2591.	86.48	1.6062E+00
U-237	2349.	59.54	1.5224E+00
NP-238	40.	984.45	2.6878E-01
NP-239	2637.	106.12	6.7593E-01
PU-239	2939.	129.30	2.2125E+03
AM-241	2349.	59.54	1.4551E+00
AM-243	2436.	74.67	4.3914E-01

*****

* 222-S Laboratory Counting Room 21-NOV-1998 03:39:10.87 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27141
 Sample ID: S98T002739-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

NE Wright

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2869.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:27.61 sec
 Dead Time: 1.0%

Verified by:

John Pelyea 23 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 21-NOV-1998 01:07:16.51
 Decayed to: 21-NOV-1998 01:07:16.51
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.47*	114261	1.46	1322.93	1315	16	0.6		CS-137	55.4
0	968.08	41	1.59	1936.18	1928	15	65.6			

Summary of Nuclide Activity

Sample ID : S98T002739-SAM

Acquisition date : 21-NOV-1998 01:07:16

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma
			Uncorrected	Decay Corr		
			uCi/L	uCi/L	2-Sigma Error	%Error
CS-137	30.00Y	1.00	5.540E+01	5.540E+01	0.034E+01	0.61
Total Activity :			5.540E+01	5.540E+01		

Grand Total Activity :	5.540E+01	5.540E+01

Flags: "K" = Keyline not found

"E" = Manually edited

"M" = Manually accepted

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002739-SAM

Acquisition date : 21-NOV-1998 01:07:16

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	3467.	477.59	8.2390E-01
NA-24	18.	1368.55	1.7852E-02
K-40	317.	1460.75	6.3728E-01
AR-41	19.	1293.60	2.5765E-02
SC-46	48.	1120.55	2.2333E-02
CR-51	2893.	320.08	5.3285E-01
MN-54	43.	834.83	1.6696E-02
CO-56	47.	846.76	1.7549E-02
CO-58	50.	810.78	1.7604E-02
FE-59	34.	1099.25	3.2895E-02
CO-60	29.	1332.50	2.0331E-02
ZN-65	43.	1115.55	4.1543E-02
SE-75	3401.	264.66	8.4314E-02
KR-85	1809.	514.00	1.5237E+01
SR-85	1808.	514.01	6.8912E-02
Y-88	6.	1836.06	1.4039E-02
Y-91	32.	1204.67	6.5286E+00
NB-94	44.	871.09	1.7371E-02
NB-95	51.	765.78	1.6863E-02
ZR-95	41.	756.73	2.7528E-02
ZRNB-95	64.	724.18	8.1518E-02
RU-103	2225.	497.08	7.8105E-02
RURH-106	799.	621.93	1.0587E+00
AG-108m	56.	722.94	1.8440E-02
CD-109	2950.	88.03	1.4260E+00
AG-110M	3083.	657.76	1.1337E-01
SN-113	3221.	391.69	1.0663E-01
TE-123m	3438.	159.00	4.4860E-02
SB-124	753.	602.73	5.0769E-02
SB-125	3741.	427.89	2.7156E-01
TE-125m	3008.	109.27	1.4043E+01
I-129	2787.	39.60	1.5957E+01
I-131	3099.	364.48	7.7393E-02
XE-131m	3491.	163.93	1.9534E+00
BA-133	3032.	356.02	9.7580E-02
CS-134	765.	604.70	5.1380E-02
CS-136	49.	818.51	1.7548E-02
CS-138	12.	1435.86	6.5764E-02
CE-139	3537.	165.85	4.8228E-02
BA-140	1276.	537.31	2.3839E-01
LA-140	12.	1596.21	1.7594E-02
CE-141	3425.	145.44	7.7700E-02
CE-144	3573.	133.51	3.4886E-01
CEPR-144	3573.	133.51	6.9710E-01
EU-152	3400.	121.78	1.3763E-01
EU-154	36.	1274.51	6.3065E-02
EU-155	3035.	86.54	1.6394E-01
HF-181	2781.	482.18	9.5844E-02
TA-182	2752.	67.75	2.2158E-01
HG-203	3245.	279.20	6.1873E-02
BI-207	868.	569.70	5.1743E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002739-SAM

Acquisition date : 21-NOV-1998 01:07:16

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	3236.	277.36	7.9269E-01
PB-210	2675.	46.50	1.0802E+01
BI-212	66.	727.18	2.7339E-01
PB-212	3821.	238.63	1.1060E-01
BI-214	808.	609.31	1.1578E-01
PB-214	3001.	351.92	6.6531E-01
RA-224	3645.	240.99	1.2073E+00
RA-226	4265.	186.10	1.2550E+00
AC-228	76.	911.21	8.7845E-02
TH-228	2976.	84.37	4.4058E+00
TH-229	2946.	88.47	2.0654E-01
U-232	2675.	57.78	7.8171E+01
PA-233	3025.	312.17	1.3909E-01
UTH-233	3665.	245.34	4.3834E+01
PA-234M	40.	1001.03	2.8650E+00
TH-234	2689.	63.29	2.9515E+00
U-235	4225.	185.71	7.5964E-02
NP-237	3038.	86.48	4.3441E-01
U-237	2765.	59.54	4.1254E-01
NP-238	32.	984.45	6.0822E-02
NP-239	3166.	106.12	1.8498E-01
PU-239	3428.	129.30	5.9694E+02
AM-241	2766.	59.54	3.9430E-01
AM-243	2844.	74.67	1.1852E-01

*****

* 222-S Laboratory Counting Room 21-NOV-1998 06:45:01.84 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27141
 Sample ID: S98T002739-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Em Baile

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2870.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:30.33 sec
 Dead Time: 1.0%

Verified by:

John Kelly 23 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 21-NOV-1998 04:13:03.91
 Decayed to: 21-NOV-1998 04:13:03.91
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.47*	119655	1.47	1322.94	1315	16	0.6		CS-137	58.0
0	1764.26*	15	1.28	3528.76	3523	11	98.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002739-DUP

Acquisition date : 21-NOV-1998 04:13:03

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	5.801E+01	5.801E+01	0.034E+01	0.59	
Total Activity :			5.801E+01	5.801E+01			

Grand Total Activity :	5.801E+01	5.801E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002739-DUP

Acquisition date : 21-NOV-1998 04:13:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	3666.	477.59	8.4700E-01
NA-24	19.	1368.55	1.8389E-02
K-40	326.	1460.75	6.4572E-01
AR-41	29.	1293.60	3.1090E-02
SC-46	48.	1120.55	2.2307E-02
CR-51	3096.	320.08	5.5103E-01
MN-54	44.	834.83	1.6900E-02
CO-56	40.	846.76	1.6313E-02
CO-58	43.	810.78	1.6371E-02
FE-59	21.	1099.25	2.6712E-02
CO-60	30.	1332.50	2.0647E-02
ZN-65	30.	1115.55	3.5428E-02
SE-75	3519.	264.66	8.5752E-02
KR-85	1858.	514.00	1.5437E+01
SR-85	1856.	514.01	6.9808E-02
Y-88	5.	1836.06	1.3092E-02
Y-91	26.	1204.67	5.9036E+00
NB-94	41.	871.09	1.6956E-02
NB-95	69.	765.78	1.9337E-02
ZR-95	52.	756.73	3.0728E-02
ZRNB-95	62.	724.18	8.0745E-02
RU-103	2337.	497.08	8.0030E-02
RURH-106	783.	621.93	1.0483E+00
AG-108m	64.	722.94	1.9562E-02
CD-109	3066.	88.03	1.4535E+00
AG-110M	3207.	657.76	1.1561E-01
SN-113	3375.	391.69	1.0912E-01
TE-123m	3667.	159.00	4.6316E-02
SB-124	846.	602.73	5.3748E-02
SB-125	3830.	427.89	2.7474E-01
TE-125m	3277.	109.27	1.4652E+01
I-129	2899.	39.60	1.6270E+01
I-131	3206.	364.48	7.8702E-02
XE-131m	3593.	163.93	1.9815E+00
BA-133	3071.	356.02	9.8195E-02
CS-134	835.	604.70	5.3643E-02
CS-136	45.	818.51	1.6766E-02
CS-138	16.	1435.86	7.4720E-02
CE-139	3612.	165.85	4.8732E-02
BA-140	1383.	537.31	2.4803E-01
LA-140	16.	1596.21	1.9943E-02
CE-141	3561.	145.44	7.9221E-02
CE-144	3667.	133.51	3.5337E-01
CEPR-144	3667.	133.51	7.0613E-01
EU-152	3555.	121.78	1.4071E-01
EU-154	39.	1274.51	6.4872E-02
EU-155	3125.	86.54	1.6632E-01
HF-181	2908.	482.18	9.7993E-02
TA-182	2879.	67.75	2.2658E-01
HG-203	3312.	279.20	6.2499E-02
BI-207	931.	569.70	5.3540E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002739-DUP

Acquisition date : 21-NOV-1998 04:13:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	3379.	277.36	8.0974E-01
PB-210	2854.	46.50	1.1153E+01
BI-212	51.	727.18	2.4124E-01
PB-212	3999.	238.63	1.1312E-01
BI-214	907.	609.31	1.2251E-01
PB-214	3096.	351.92	6.7583E-01
RA-224	3849.	240.99	1.2402E+00
RA-226	4597.	186.10	1.3025E+00
AC-228	75.	911.21	8.7121E-02
TH-228	3132.	84.37	4.5185E+00
TH-229	3063.	88.47	2.1057E-01
U-232	2780.	57.78	7.9686E+01
PA-233	3128.	312.17	1.4141E-01
UTH-233	3900.	245.34	4.5202E+01
PA-234M	35.	1001.03	2.7091E+00
TH-234	2901.	63.29	3.0642E+00
U-235	4514.	185.71	7.8496E-02
NP-237	3133.	86.48	4.4110E-01
U-237	2852.	59.54	4.1891E-01
NP-238	39.	984.45	6.6552E-02
NP-239	3236.	106.12	1.8699E-01
PU-239	3707.	129.30	6.2046E+02
AM-241	2852.	59.54	4.0038E-01
AM-243	2913.	74.67	1.1993E-01

*****

* 222-S Laboratory Counting Room 21-NOV-1998 09:44:44.60 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27141
 Sample ID: S98T2844-SAM
 Sample Size: 5.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

Em Barlow

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2871.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:35.67 sec
 Dead Time: 1.1%

Verified by:

John Rebyea 23 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 21-NOV-1998 07:12:39.38
 Decayed to: 21-NOV-1998 07:12:39.38
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.48*	134237	1.48	1322.96	1315	16	0.6		CS-137	130.
0	968.91*	23	1.00	1937.83	1932	111	19.9			
0	1764.07*	12	2.21	3528.37	3522	121	19.9			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2844-SAM

Acquisition date : 21-NOV-1998 07:12:39

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.302E+02	1.302E+02	0.007E+02	0.56	
Total Activity :			1.302E+02	1.302E+02			

Grand Total Activity :	1.302E+02	1.302E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2844-SAM

Acquisition date : 21-NOV-1998 07:12:39

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	4077.	477.59	1.7856E+00
NA-24	16.	1368.55	3.4284E-02
K-40	317.	1460.75	1.2750E+00
AR-41	21.	1293.60	5.3880E-02
SC-46	55.	1120.55	4.7366E-02
CR-51	3518.	320.08	1.1740E+00
MN-54	34.	834.83	2.9965E-02
CO-56	57.	846.76	3.8448E-02
CO-58	63.	810.78	3.9160E-02
FE-59	29.	1099.25	6.1093E-02
CO-60	37.	1332.50	4.5446E-02
ZN-65	41.	1115.55	8.1230E-02
SE-75	3961.	264.66	1.8187E-01
KR-85	2122.	514.00	3.2972E+01
SR-85	2122.	514.01	1.4915E-01
Y-88	6.	1836.06	2.8296E-02
Y-91	37.	1204.67	1.3883E+01
NB-94	49.	871.09	3.6551E-02
NB-95	44.	765.78	3.1437E-02
ZR-95	58.	756.73	6.4993E-02
ZRNB-95	67.	724.18	1.6683E-01
RU-103	2450.	497.08	1.6386E-01
RURH-106	856.	621.93	2.1904E+00
AG-108m	61.	722.94	3.8378E-02
CD-109	3567.	88.03	3.1330E+00
AG-110M	3519.	657.76	2.4208E-01
SN-113	3794.	391.69	2.3126E-01
TE-123m	4081.	159.00	9.7671E-02
SB-124	936.	602.73	1.1295E-01
SB-125	4281.	427.89	5.8065E-01
TE-125m	3765.	109.27	3.1391E+01
I-129	3457.	39.60	3.5502E+01
I-131	3572.	364.48	1.6605E-01
XE-131m	4030.	163.93	4.1949E+00
BA-133	3466.	356.02	2.0851E-01
CS-134	899.	604.70	1.1123E-01
CS-136	53.	818.51	3.6287E-02
CS-138	13.	1435.86	1.3652E-01
CE-139	4099.	165.85	1.0376E-01
BA-140	1527.	537.31	5.2085E-01
LA-140	8.	1596.21	2.8876E-02
CE-141	4022.	145.44	1.6827E-01
CE-144	4115.	133.51	7.4832E-01
CEPR-144	4115.	133.51	1.4952E+00
EU-152	4031.	121.78	2.9948E-01
EU-154	27.	1274.51	1.1005E-01
EU-155	3542.	86.54	3.5393E-01
HF-181	3318.	482.18	2.0919E-01
TA-182	3415.	67.75	4.9312E-01
HG-203	3717.	279.20	1.3235E-01
BI-207	1058.	569.70	1.1399E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2844-SAM

Acquisition date : 21-NOV-1998 07:12:39

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	3812.	277.36	1.7192E+00
PB-210	3195.	46.50	2.3588E+01
BI-212	64.	727.18	5.3559E-01
PB-212	4317.	238.63	2.3498E-01
BI-214	978.	609.31	2.5422E-01
PB-214	3685.	351.92	1.4741E+00
RA-224	4403.	240.99	2.6513E+00
RA-226	5106.	186.10	2.7442E+00
AC-228	68.	911.21	1.6615E-01
TH-228	3607.	84.37	9.6922E+00
TH-229	3590.	88.47	4.5553E-01
U-232	3123.	57.78	1.6880E+02
PA-233	3464.	312.17	2.9747E-01
UTH-233	4181.	245.34	9.3575E+01
PA-234M	36.	1001.03	5.4526E+00
TH-234	3369.	63.29	6.5989E+00
U-235	5065.	185.71	1.6623E-01
NP-237	3545.	86.48	9.3780E-01
U-237	3189.	59.54	8.8543E-01
NP-238	33.	984.45	1.2410E-01
NP-239	3741.	106.12	4.0183E-01
PU-239	4113.	129.30	1.3065E+03
AM-241	3190.	59.54	8.4628E-01
AM-243	3273.	74.67	2.5412E-01

*****
 * 222-S Laboratory Counting Room 21-NOV-1998 12:20:17.81 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27141
 Sample ID: S98T2844-DUP
 Sample Size: 5.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

Em Barlow

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2872.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:35.51 sec
 Dead Time: 1.1%

Verified by:

John Relyea 23 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 21-NOV-1998 09:48:13.62
 Decayed to: 21-NOV-1998 09:48:13.62
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.50*	142617	1.47	1323.00	1315	16	0.5		CS-137	138.
0	1460.50*	28	2.25	2921.12	2912	17156.8			K-40	0.419

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2844-DUP

Acquisition date : 21-NOV-1998 09:48:13

Total number of lines in spectrum	2
Number of unidentified lines	0
Number of lines tentatively identified by NID	2 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
K-40	1.28E+09Y	1.00	4.186E-01	4.186E-01	6.564E-01	156.79	
CS-137	30.00Y	1.00	1.383E+02	1.383E+02	0.008E+02	0.54	
Total Activity :			1.387E+02	1.387E+02			

Grand Total Activity : 1.387E+02   1.387E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2844-DUP

Acquisition date : 21-NOV-1998 09:48:13

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	4225.	477.59	1.8174E+00
NA-24	17.	1368.55	3.4863E-02
AR-41	32.	1293.60	6.4925E-02
SC-46	69.	1120.55	5.2568E-02
CR-51	3593.	320.08	1.1864E+00
MN-54	38.	834.83	3.1376E-02
CO-56	42.	846.76	3.3275E-02
CO-58	50.	810.78	3.4991E-02
FE-59	30.	1099.25	6.2813E-02
CO-60	24.	1332.50	3.7287E-02
ZN-65	36.	1115.55	7.6337E-02
SE-75	3998.	264.66	1.8270E-01
KR-85	2274.	514.00	3.4113E+01
SR-85	2273.	514.01	1.5428E-01
Y-88	11.	1836.06	3.6542E-02
Y-91	24.	1204.67	1.1558E+01
NB-94	40.	871.09	3.3520E-02
NB-95	47.	765.78	3.2448E-02
ZR-95	61.	756.73	6.6503E-02
ZRNB-95	60.	724.18	1.5854E-01
RU-103	2711.	497.08	1.7226E-01
RURH-106	953.	621.93	2.3092E+00
AG-108m	56.	722.94	3.6874E-02
CD-109	3813.	88.03	3.2383E+00
AG-110M	3557.	657.76	2.4337E-01
SN-113	3978.	391.69	2.3676E-01
TE-123m	4216.	159.00	9.9261E-02
SB-124	968.	602.73	1.1485E-01
SB-125	4488.	427.89	5.9438E-01
TE-125m	3926.	109.27	3.2046E+01
I-129	3433.	39.60	3.5382E+01
I-131	3647.	364.48	1.6777E-01
XE-131m	4335.	163.93	4.3490E+00
BA-133	3682.	356.02	2.1484E-01
CS-134	980.	604.70	1.1606E-01
CS-136	57.	818.51	3.7679E-02
CS-138	13.	1435.86	1.3458E-01
CE-139	4226.	165.85	1.0535E-01
BA-140	1555.	537.31	5.2555E-01
LA-140	18.	1596.21	4.1066E-02
CE-141	4364.	145.44	1.7522E-01
CE-144	4391.	133.51	7.7276E-01
CEPR-144	4391.	133.51	1.5441E+00
EU-152	4203.	121.78	3.0576E-01
EU-154	31.	1274.51	1.1812E-01
EU-155	3758.	86.54	3.6446E-01
HF-181	3491.	482.18	2.1452E-01
TA-182	3456.	67.75	4.9603E-01
HG-203	3920.	279.20	1.3589E-01
BI-207	1095.	569.70	1.1594E-01
TL-208	4112.	277.36	1.7850E+00

Minimum Detectable Activity Report (continued)

HNF-1663 REV. 0

Page : 4

Sample ID : S98T2844-DUP

Acquisition date : 21-NOV-1998 09:48:13

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	3482.	46.50	2.4615E+01
BI-212	64.	727.18	5.3578E-01
PB-212	4657.	238.63	2.4399E-01
BI-214	1047.	609.31	2.6284E-01
PB-214	3763.	351.92	1.4896E+00
RA-224	4680.	240.99	2.7328E+00
RA-226	5324.	186.10	2.8016E+00
AC-228	71.	911.21	1.6976E-01
TH-228	3789.	84.37	9.9309E+00
TH-229	3834.	88.47	4.7060E-01
U-232	3425.	57.78	1.7669E+02
PA-233	3805.	312.17	3.1164E-01
UTH-233	4556.	245.34	9.7642E+01
PA-234M	36.	1001.03	5.4480E+00
TH-234	3415.	63.29	6.6440E+00
U-235	5258.	185.71	1.6934E-01
NP-237	3761.	86.48	9.6569E-01
U-237	3480.	59.54	9.2448E-01
NP-238	34.	984.45	1.2471E-01
NP-239	4015.	106.12	4.1616E-01
PU-239	4431.	129.30	1.3557E+03
AM-241	3480.	59.54	8.8358E-01
AM-243	3629.	74.67	2.6742E-01

LABCORE Completed Worklist Report for Worklist# 27142

Analyst: gl Instrument: GEA03

Book#: 28B57

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105, @GEA-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@GEA-01 CO60-02 SOLID		4.513e-03	4.76e-03	105.473 % Recovery	
1 STD	0		@GEA-01 CO60-02E SOLID		1	2.75	2.750 % Ct Error	
1 STD	0		@GEA-01 CS13702 SOLID		7.448e-03	7.43e-03	99.758 % Recovery	
1 STD	0		@GEA-01 CS13702E SOLID		1	2.71	2.710 % Ct Error	
2 BLNK-PREP	0		@GEA-01 CO60-02 SOLID		1	<9.30e-3		uCi/g
2 BLNK-PREP	0		@GEA-01 CS13702 SOLID		1	<1.80e-2		uCi/g
3 SAMPLE	S98T002920	0 F	@GEA-01 CO60-02 SOLID		N/A	8.421e-03	8.42e-003	uCi/g
3 SAMPLE	S98T002920	0 F	@GEA-01 CO60-02E SOLID		N/A	n/a		% Ct. Error
3 SAMPLE	S98T002920	0 F	@GEA-01 CS13702 SOLID		N/A	7.543e+00		uCi/g
3 SAMPLE	S98T002920	0 F	@GEA-01 CS13702E SOLID		N/A	1.02		% Ct. Error
4 DUP	S98T002920	0 F	@GEA-01 CO60-02 SOLID		<8.42e-3	<8.35e-3		RPD
4 DUP	S98T002920	0 F	@GEA-01 CO60-02E SOLID		1	n/a		% Ct Error
4 DUP	S98T002920	0 F	@GEA-01 CS13702 SOLID		7.54e+00	7.89e+00	4.537 RPD	
4 DUP	S98T002920	0 F	@GEA-01 CS13702E SOLID		1.00	0.990	0.990 % Ct Error	
5 SAMPLE	S98T002921	0 F	@GEA-01 CO60-02 SOLID		N/A	8.425e-03	8.43e-003	uCi/g
5 SAMPLE	S98T002921	0 F	@GEA-01 CO60-02E SOLID		N/A	n/a		% Ct. Error
5 SAMPLE	S98T002921	0 F	@GEA-01 CS13702 SOLID		N/A	3.332e+01		uCi/g
5 SAMPLE	S98T002921	0 F	@GEA-01 CS13702E SOLID		N/A	0.480		% Ct. Error
6 DUP	S98T002921	0 F	@GEA-01 CO60-02 SOLID		<8.43e-3	<8.05e-3		RPD
6 DUP	S98T002921	0 F	@GEA-01 CO60-02E SOLID		1	n/a		% Ct Error
6 DUP	S98T002921	0 F	@GEA-01 CS13702 SOLID		3.33e+01	3.37e+01	1.194 RPD	
6 DUP	S98T002921	0 F	@GEA-01 CS13702E SOLID		1.00	0.480	0.480 % Ct Error	
7 SAMPLE	S98T002922	0 F	@GEA-01 CO60-02 SOLID		N/A	8.425e-02	8.43e-002	uCi/g
7 SAMPLE	S98T002922	0 F	@GEA-01 CO60-02E SOLID		N/A	n/a		% Ct. Error
7 SAMPLE	S98T002922	0 F	@GEA-01 CS13702 SOLID		N/A	7.038e+01		uCi/g
7 SAMPLE	S98T002922	0 F	@GEA-01 CS13702E SOLID		N/A	1.08		% Ct. Error
8 DUP	S98T002922	0 F	@GEA-01 CO60-02 SOLID		<8.43e-2	<8.84e-2		RPD
8 DUP	S98T002922	0 F	@GEA-01 CO60-02E SOLID		1	n/a		% Ct Error
8 DUP	S98T002922	0 F	@GEA-01 CS13702 SOLID		7.04e+01	6.15e+01	13.495 RPD	
8 DUP	S98T002922	0 F	@GEA-01 CS13702E SOLID		1.00	1.15	1.150 % Ct Error	

Comments Section:

Comments for sample# S98T002920 and test @GEA-01 .
DL=0 => n/a

Comments for sample# S98T002921 and test @GEA-01 .
DL=0 => n/a

Comments for sample# S98T002922 and test @GEA-01 .

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27142

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
		DL=0	=	>	n/a.					

Final page for worklist# 27142

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date 11-24-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27142

Analyst: gll Instrument: GEA00 #3 Book# 28B57

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105, @GEA-01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	SOLID		
2 BLNK-PREP			@GEA-01	SOLID		
3 SAMPLE	S98T002920	0 F	@GEA-01	SOLID	98000490	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4 DUP	S98T002920	0 F	@GEA-01	SOLID		
5 SAMPLE	S98T002921	0 F	@GEA-01	SOLID	98000490	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6 DUP	S98T002921	0 F	@GEA-01	SOLID		
7 SAMPLE	S98T002922	0 F	@GEA-01	SOLID	98000490	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8 DUP	S98T002922	0 F	@GEA-01	SOLID		

Final page for worklist # 27142

gll 11-23-98
Analyst Signature Date

Sybil G. Chone 11/24/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 12:00:50.41 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27142
 Sample ID: WL27142-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3913.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:02.79 sec
 Dead Time: 0.1%

Verified by:

RF Jones 11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 11:10:14.83
 Decayed to: 23-NOV-1998 11:10:14.83
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.71*	6570	1.63	1323.60	1315	17	2.7		CS-137	7.43
0	1173.18	3082	1.89	2346.51	2340	14	3.8		CO-60	4.86
0	1332.37	2660	1.94	2664.96	2654	19	4.0		CO-60	4.66

Summary of Nuclide Activity

Page : 2

Sample ID : WL27142-STD

Acquisition date : 23-NOV-1998 11:10:14

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.759E+00	4.759E+00	0.131E+00	2.75	
CS-137	30.00Y	1.00	7.425E+00	7.425E+00	0.201E+00	2.71	
Total Activity :			1.218E+01	1.218E+01			

Grand Total Activity :	1.218E+01	1.218E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : WL27142-STD

Acquisition date : 23-NOV-1998 11:10:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	341.	477.59	6.2787E-01
NA-24	12.	1368.55	3.3850E-02
K-40	123.	1460.75	9.5688E-01
AR-41	31.	1293.60	5.7742E-02
SC-46	136.	1120.55	8.6320E-02
CR-51	362.	320.08	4.8618E-01
MN-54	144.	834.83	6.8822E-02
CO-56	177.	846.76	7.6884E-02
CO-58	150.	810.78	6.8723E-02
FE-59	160.	1099.25	1.6211E-01
ZN-65	145.	1115.55	1.7465E-01
SE-75	433.	264.66	7.9489E-02
KR-85	248.	514.00	1.3597E+01
SR-85	248.	514.01	6.1497E-02
Y-88	5.	1836.06	2.9463E-02
Y-91	61.	1204.67	2.0887E+01
NB-94	177.	871.09	7.8844E-02
NB-95	135.	765.78	6.2136E-02
ZR-95	139.	756.73	1.1401E-01
ZRNB-95	120.	724.18	2.5648E-01
RU-103	239.	497.08	6.2009E-02
RURH-106	171.	621.93	1.1717E+00
AG-108m	124.	722.94	6.2263E-02
CD-109	405.	88.03	1.1868E+00
AG-110M	297.	657.76	8.3941E-02
SN-113	351.	391.69	8.7767E-02
TE-123m	433.	159.00	3.8069E-02
SB-124	145.	602.73	5.3509E-02
SB-125	363.	427.89	2.0830E-01
TE-125m	365.	109.27	1.1099E+01
I-129	430.	39.60	0.0000E+00
I-131	333.	364.48	6.4047E-02
XE-131m	384.	163.93	1.5630E+00
BA-133	328.	356.02	8.1588E-02
CS-134	161.	604.70	5.6387E-02
CS-136	164.	818.51	7.2164E-02
CS-138	10.	1435.86	6.7622E-02
CE-139	395.	165.85	3.9066E-02
BA-140	188.	537.31	2.2002E-01
LA-140	8.	1596.21	3.3411E-02
CE-141	428.	145.44	6.4294E-02
CE-144	414.	133.51	2.7420E-01
CEPR-144	414.	133.51	5.4780E-01
EU-152	382.	121.78	1.0549E-01
EU-154	25.	1274.51	1.2769E-01
EU-155	388.	86.54	1.3180E-01
HF-181	300.	482.18	7.6423E-02
TA-182	332.	67.75	1.8361E-01
HG-203	363.	279.20	5.4398E-02
BI-207	188.	569.70	5.7609E-02
TL-208	361.	277.36	6.9627E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27142-STD

Acquisition date : 23-NOV-1998 11:10:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	418.	46.50	0.0000E+00
BI-212	139.	727.18	9.0253E-01
PB-212	476.	238.63	1.0342E-01
BI-214	195.	609.31	1.3558E-01
PB-214	342.	351.92	2.5516E-01
RA-224	496.	240.99	1.1720E+00
RA-226	495.	186.10	1.0690E+00
AC-228	220.	911.21	3.4176E-01
TH-228	348.	84.37	3.3968E+00
TH-229	407.	88.47	1.7248E-01
U-232	440.	57.78	1.0086E+02
PA-233	365.	312.17	1.2505E-01
UTH-233	419.	245.34	3.9345E+01
PA-234M	209.	1001.03	1.4767E+01
TH-234	326.	63.29	2.6240E+00
U-235	505.	185.71	6.5601E-02
NP-237	389.	86.48	3.4927E-01
U-237	475.	59.54	4.8944E-01
NP-238	178.	984.45	3.1766E-01
NP-239	340.	106.12	1.3598E-01
PU-239	389.	129.30	4.6231E+02
AM-241	475.	59.54	4.6943E-01
AM-243	365.	74.67	9.7005E-02

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 14:44:32.96 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27142
 Sample ID: WL27142-BLNK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Sh Hake

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3915.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:02.25 sec
 Dead Time: 0.0%

Verified by:

RA Jones 11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 12:13:57.88
 Decayed to: 23-NOV-1998 12:13:57.88
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	352.17*	134	1.48	704.73	700	10 40.2			
0	609.50*	118	1.61	1219.21	1213	13 40.5			
0	1460.59*	44	2.15	2921.48	2914	16 106.3		K-40	0.261

Summary of Nuclide Activity

Page : 2

Sample ID : WL27142-BLANK

Acquisition date : 23-NOV-1998 12:13:57

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~2.608E-01~~	~~2.608E-01~~	~~2.773E-01~~	~~106.33~~	~~11-24-98~~
Total Activity :			2.608E-01	2.608E-01			

Grand Total Activity : 2.608E-01   2.608E-01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27142-BLNK

Acquisition date : 23-NOV-1998 12:13:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	108.	477.59	1.2037E-01
NA-24	28.	1368.55	1.7244E-02
AR-41	27.	1293.60	2.3657E-02
SC-46	70.	1120.55	2.1072E-02
CR-51	173.	320.08	1.1359E-01
MN-54	54.	834.83	1.4505E-02
CO-56	56.	846.76	1.4849E-02
CO-58	62.	810.78	1.5144E-02
FE-59	53.	1099.25	3.2221E-02
CO-60	41.	1332.50	1.8984E-02
ZN-65	39.	1115.55	3.1658E-02
SE-75	202.	264.66	1.8331E-02
KR-85	143.	514.00	3.4831E+00
SR-85	143.	514.01	1.5728E-02
Y-88	13.	1836.06	1.4533E-02
Y-91	47.	1204.67	6.1991E+00
NB-94	40.	871.09	1.3090E-02
NB-95	52.	765.78	1.3240E-02
ZR-95	53.	756.73	2.4193E-02
ZRNB-95	58.	724.18	6.0856E-02
RU-103	104.	497.08	1.3889E-02
RURH-106	72.	621.93	2.5825E-01
AG-108m	59.	722.94	1.4616E-02
CD-109	240.	88.03	3.0718E-01
AG-110M	78.	657.76	1.4759E-02
SN-113	152.	391.69	1.9554E-02
TE-123m	256.	159.00	9.8358E-03
SB-124	78.	602.73	1.3298E-02
SB-125	118.	427.89	4.0474E-02
TE-125m	253.	109.27	3.0957E+00
I-129	406.	39.60	0.0000E+00
I-131	147.	364.48	1.4434E-02
XE-131m	286.	163.93	4.5274E-01
BA-133	136.	356.02	1.7811E-02
CS-134	80.	604.70	1.3512E-02
CS-136	46.	818.51	1.3313E-02
CS-137	415.	661.66	3.6717E-02
CS-138	18.	1435.86	6.1210E-02
CE-139	265.	165.85	1.0730E-02
BA-140	87.	537.31	5.0891E-02
LA-140	16.	1596.21	1.5407E-02
CE-141	275.	145.44	1.7295E-02
CE-144	255.	133.51	7.2339E-02
CEPR-144	256.	133.51	1.4458E-01
EU-152	272.	121.78	2.9814E-02
EU-154	38.	1274.51	5.0973E-02
EU-155	287.	86.54	3.7936E-02
HF-181	124.	482.18	1.6676E-02
TA-182	226.	67.75	5.0766E-02
HG-203	193.	279.20	1.3356E-02
BI-207	92.	569.70	1.3642E-02

. 4460

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27142-BLANK

Acquisition date : 23-NOV-1998 12:13:57

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	202.	277.36	1.7547E-01
PB-210	472.	46.50	0.0000E+00
BI-212	68.	727.18	2.1516E-01
PB-212	296.	238.63	2.7400E-02
BI-214	242.	609.31	5.0123E-02
PB-214	315.	351.92	1.8149E-01
RA-224	277.	240.99	2.9627E-01
RA-226	278.	186.10	2.6912E-01
AC-228	78.	911.21	6.9341E-02
TH-228	264.	84.37	9.9038E-01
TH-229	233.	88.47	4.3907E-02
U-232	408.	57.78	3.2413E+01
PA-233	170.	312.17	2.8845E-02
UTH-233	240.	245.34	1.0022E+01
PA-234M	49.	1001.03	2.4793E+00
TH-234	210.	63.29	7.0797E-01
U-235	274.	185.71	1.6262E-02
NP-237	289.	86.48	1.0088E-01
U-237	361.	59.54	1.4321E-01
NP-238	40.	984.45	5.2811E-02
NP-239	232.	106.12	3.8099E-02
PU-239	286.	129.30	1.3283E+02
AM-241	361.	59.54	1.3687E-01
AM-243	270.	74.67	2.7975E-02

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 17:24:57.67 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27142
 Sample ID: S98T2920-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Shirley L. Wade

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3916.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.33 sec
 Dead Time: 0.1%

Verified by:

R. Jones 11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-NOV-1998 14:54:17.39
 Decayed to: 23-NOV-1998 14:54:17.39
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	609.38*	58	1.54	1218.96	1215	8107.8			
0	661.73*	40967	1.60	1323.63	1315	17 1.0		CS-137	15.4
0	1460.65*	27	1.98	2921.61	2914	16178.0		K-40	0.156

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2920-SAM

Acquisition date : 23-NOV-1998 14:54:17

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma Error	2-Sigma	Flags
			Uncorrected uCi/L	Decay Corr uCi/L				
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.560E-01~~	~~1.560E-01~~	~~2.775E-01~~	~~177.96~~		~~PA~~
CS-137	30.00Y	1.00	1.543E+01	1.543E+01	0.016E+01	1.02		
Total Activity :			1.559E+01	1.559E+01				11-24-98

Grand Total Activity :	1.559E+01	1.559E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2920-SAM

Acquisition date : 23-NOV-1998 14:54:17

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1459.	477.59	4.2599E-01
NA-24	18.	1368.55	1.4062E-02
AR-41	37.	1293.60	2.7505E-02
SC-46	91.	1120.55	2.3721E-02
CR-51	1161.	320.08	2.8665E-01
MN-54	50.	834.83	1.3892E-02
CO-56	50.	846.76	1.4159E-02
CO-58	52.	810.78	1.3916E-02
FE-59	38.	1099.25	2.7520E-02
CO-60	33.	1332.50	1.7192E-02
ZN-65	35.	1115.55	3.0105E-02
SE-75	1367.	264.66	4.6506E-02
KR-85	777.	514.00	7.9012E+00
SR-85	776.	514.01	3.5726E-02
Y-88	7.	1836.06	1.1288E-02
Y-91	44.	1204.67	6.0381E+00
NB-94	63.	871.09	1.6067E-02
NB-95	74.	765.78	1.5593E-02
ZR-95	72.	756.73	2.7891E-02
ZRNB-95	60.	724.18	6.1747E-02
RU-103	877.	497.08	3.8953E-02
RURH-106	299.	621.93	5.1040E-01
AG-108m	61.	722.94	1.4905E-02
CD-109	1228.	88.03	6.8050E-01
AG-110M	1024.	657.76	5.1161E-02
SN-113	1252.	391.69	5.4511E-02
TE-123m	1447.	159.00	2.2926E-02
SB-124	294.	602.73	2.5023E-02
SB-125	1421.	427.89	1.3541E-01
TE-125m	1288.	109.27	6.8502E+00
I-129	1330.	39.60	0.0000E+00
I-131	1215.	364.48	4.0278E-02
XE-131m	1447.	163.93	9.9897E-01
BA-133	1197.	356.02	5.1190E-02
CS-134	327.	604.70	2.6473E-02
CS-136	44.	818.51	1.3006E-02
CS-138	14.	1435.86	5.3961E-02
CE-139	1441.	165.85	2.4534E-02
BA-140	534.	537.31	1.2178E-01
LA-140	11.	1596.21	1.3218E-02
CE-141	1382.	145.44	3.8073E-02
CE-144	1416.	133.51	1.6686E-01
CEPR-144	1416.	133.51	3.3343E-01
EU-152	1393.	121.78	6.6190E-02
EU-154	22.	1274.51	4.0307E-02
EU-155	1242.	86.54	7.7600E-02
HF-181	1186.	482.18	4.9857E-02
TA-182	1120.	67.75	1.1078E-01
HG-203	1322.	279.20	3.4117E-02
BI-207	332.	569.70	2.5271E-02
TL-208	1306.	277.36	4.3553E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2920-SAM

Acquisition date : 23-NOV-1998 14:54:17

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1358.	46.50	0.0000E+00
BI-212	71.	727.18	2.1965E-01
PB-212	1580.	238.63	6.2104E-02
BI-214	450.	609.31	6.7716E-02
PB-214	1325.	351.92	3.6646E-01
RA-224	1582.	240.99	6.9439E-01
RA-226	1770.	186.10	6.6577E-01
AC-228	111.	911.21	8.2180E-02
TH-228	1207.	84.37	2.0785E+00
TH-229	1231.	88.47	9.8727E-02
U-232	1316.	57.78	5.7509E+01
PA-233	1169.	312.17	7.3692E-02
UTH-233	1499.	245.34	2.4496E+01
PA-234M	48.	1001.03	2.4589E+00
TH-234	1160.	63.29	1.6257E+00
U-235	1753.	185.71	4.0269E-02
NP-237	1240.	86.48	2.0535E-01
U-237	1260.	59.54	2.6395E-01
NP-238	50.	984.45	5.8872E-02
NP-239	1264.	106.12	8.6988E-02
PU-239	1418.	129.30	2.9039E+02
AM-241	1260.	59.54	2.5227E-01
AM-243	1255.	74.67	5.9186E-02

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 19:57:48.42 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27142
 Sample ID: S98T2920-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3917.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.57 sec
 Dead Time: 0.1%

Verified by:

RAJmer 11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 17:27:08.55
 Decayed to: 23-NOV-1998 17:27:08.55
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.73*	42803	1.60	1323.64	1315	18	1.0		CS-137	16.1

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2920-DUP

Acquisition date : 23-NOV-1998 17:27:08

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected	Decay Corr	2-Sigma Error	%Error	
			uCi/L	uCi/L			
CS-137	30.00Y	1.00	1.613E+01	1.613E+01	0.016E+01	0.99	
Total Activity :			1.613E+01	1.613E+01			

Grand Total Activity :	1.613E+01	1.613E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2920-DUP

Acquisition date : 23-NOV-1998 17:27:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1529.	477.59	4.3602E-01
NA-24	32.	1368.55	1.8218E-02
K-40	390.	1460.75	5.5445E-01
AR-41	41.	1293.60	2.8874E-02
SC-46	76.	1120.55	2.1905E-02
CR-51	1266.	320.08	2.9914E-01
MN-54	72.	834.83	1.6540E-02
CO-56	50.	846.76	1.4173E-02
CO-58	43.	810.78	1.2708E-02
FE-59	47.	1099.25	3.0437E-02
CO-60	32.	1332.50	1.7032E-02
ZN-65	37.	1115.55	3.0914E-02
SE-75	1359.	264.66	4.6362E-02
KR-85	859.	514.00	8.3003E+00
SR-85	858.	514.01	3.7519E-02
Y-88	9.	1836.06	1.2528E-02
Y-91	46.	1204.67	6.1444E+00
NB-94	54.	871.09	1.4955E-02
NB-95	54.	765.78	1.3473E-02
ZR-95	64.	756.73	2.6406E-02
ZRNB-95	72.	724.18	6.7239E-02
RU-103	937.	497.08	4.0246E-02
RURH-106	335.	621.93	5.3982E-01
AG-108m	72.	722.94	1.6060E-02
CD-109	1263.	88.03	6.9009E-01
AG-110M	1043.	657.76	5.1633E-02
SN-113	1299.	391.69	5.5511E-02
TE-123m	1476.	159.00	2.3153E-02
SB-124	353.	602.73	2.7339E-02
SB-125	1511.	427.89	1.3960E-01
TE-125m	1325.	109.27	6.9451E+00
I-129	1345.	39.60	0.0000E+00
I-131	1210.	364.48	4.0196E-02
XE-131m	1412.	163.93	9.8679E-01
BA-133	1254.	356.02	5.2367E-02
CS-134	344.	604.70	2.7126E-02
CS-136	53.	818.51	1.4189E-02
CS-138	21.	1435.86	6.4793E-02
CE-139	1420.	165.85	2.4357E-02
BA-140	578.	537.31	1.2658E-01
LA-140	13.	1596.21	1.3981E-02
CE-141	1459.	145.44	3.9101E-02
CE-144	1529.	133.51	1.7327E-01
CEPR-144	1529.	133.51	3.4624E-01
EU-152	1403.	121.78	6.6437E-02
EU-154	36.	1274.51	5.0055E-02
EU-155	1282.	86.54	7.8810E-02
HF-181	1183.	482.18	4.9790E-02
TA-182	1130.	67.75	1.1129E-01
HG-203	1322.	279.20	3.4122E-02
BI-207	384.	569.70	2.7134E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2920-DUP

Acquisition date : 23-NOV-1998 17:27:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1376.	277.36	4.4691E-01
PB-210	1453.	46.50	0.0000E+00
BI-212	58.	727.18	1.9994E-01
PB-212	1656.	238.63	6.3548E-02
BI-214	454.	609.31	6.8060E-02
PB-214	1307.	351.92	3.6404E-01
RA-224	1672.	240.99	7.1367E-01
RA-226	1837.	186.10	6.7805E-01
AC-228	98.	911.21	7.7263E-02
TH-228	1343.	84.37	2.1899E+00
TH-229	1273.	88.47	1.0039E-01
U-232	1343.	57.78	5.8092E+01
PA-233	1210.	312.17	7.4939E-02
UTH-233	1516.	245.34	2.4635E+01
PA-234M	54.	1001.03	2.5990E+00
TH-234	1198.	63.29	1.6515E+00
U-235	1829.	185.71	4.1120E-02
NP-237	1287.	86.48	2.0913E-01
U-237	1322.	59.54	2.7029E-01
NP-238	43.	984.45	5.5110E-02
NP-239	1289.	106.12	8.7840E-02
PU-239	1500.	129.30	2.9851E+02
AM-241	1322.	59.54	2.5833E-01
AM-243	1290.	74.67	5.9987E-02

*****
 * 222-S Laboratory Counting Room 23-NOV-1998 22:48:11.18 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27142
 Sample ID: S98T002921-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3918.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:24.44 sec
 Dead Time: 0.3%

Verified by:

R. Jones 11-24-98

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-NOV-1998 20:17:14.17
 Decayed to: 23-NOV-1998 20:17:14.17
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.68*	180203	1.59	1323.53	1315	18	0.5		CS-137	67.9

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002921-SAM

Acquisition date : 23-NOV-1998 20:17:14

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	6.789E+01	6.789E+01	0.033E+01	0.48	
Total Activity :			6.789E+01	6.789E+01			

Grand Total Activity :	6.789E+01	6.789E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002921-SAM

Acquisition date : 23-NOV-1998 20:17:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	5860.	477.59	8.4742E-01
NA-24	11.	1368.55	1.1462E-02
K-40	362.	1460.75	5.3464E-01
AR-41	29.	1293.60	2.4470E-02
SC-46	74.	1120.55	2.1527E-02
CR-51	4619.	320.08	5.6696E-01
MN-54	58.	834.83	1.4967E-02
CO-56	64.	846.76	1.5867E-02
CO-58	53.	810.78	1.4112E-02
FE-59	46.	1099.25	3.0160E-02
CO-60	33.	1332.50	1.7170E-02
ZN-65	49.	1115.55	3.5119E-02
SE-75	5310.	264.66	9.0953E-02
KR-85	2897.	514.00	1.5106E+01
SR-85	2896.	514.01	6.8320E-02
Y-88	8.	1836.06	1.1902E-02
Y-91	42.	1204.67	5.8879E+00
NB-94	49.	871.09	1.4286E-02
NB-95	73.	765.78	1.5495E-02
ZR-95	54.	756.73	2.4442E-02
ZRNB-95	69.	724.18	6.5855E-02
RU-103	3587.	497.08	7.8024E-02
RURH-106	1125.	621.93	9.7485E-01
AG-108m	57.	722.94	1.4425E-02
CD-109	4693.	88.03	1.3197E+00
AG-110M	4800.	657.76	1.0971E-01
SN-113	5042.	391.69	1.0851E-01
TE-123m	5283.	159.00	4.3489E-02
SB-124	1159.	602.73	4.8869E-02
SB-125	5859.	427.89	2.7287E-01
TE-125m	4670.	109.27	1.2945E+01
I-129	4503.	39.60	0.0000E+00
I-131	4718.	364.48	7.8721E-02
XE-131m	5284.	163.93	1.8952E+00
BA-133	4761.	356.02	1.0123E-01
CS-134	1148.	604.70	4.8860E-02
CS-136	63.	818.51	1.5339E-02
CS-138	23.	1435.86	6.7597E-02
CE-139	5278.	165.85	4.6616E-02
BA-140	2072.	537.31	2.3700E-01
LA-140	22.	1596.21	1.7538E-02
CE-141	5211.	145.44	7.3381E-02
CE-144	5284.	133.51	3.1997E-01
CEPR-144	5284.	133.51	6.3937E-01
EU-152	5063.	121.78	1.2529E-01
EU-154	38.	1274.51	5.1192E-02
EU-155	4560.	86.54	1.4752E-01
HF-181	4654.	482.18	9.7939E-02
TA-182	4216.	67.75	2.1318E-01
HG-203	4999.	279.20	6.5847E-02
BI-207	1310.	569.70	4.9420E-02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	5004.	277.36	8.4591E-01
PB-210	4495.	46.50	0.0000E+00
BI-212	77.	727.18	2.2810E-01
PB-212	5915.	238.63	1.1932E-01
BI-214	1226.	609.31	1.1062E-01
PB-214	4799.	351.92	6.9352E-01
RA-224	5865.	240.99	1.3279E+00
RA-226	6717.	186.10	1.2882E+00
AC-228	105.	911.21	8.0090E-02
TH-228	4471.	84.37	3.9678E+00
TH-229	4677.	88.47	1.9095E-01
U-232	4341.	57.78	1.0371E+02
PA-233	4732.	312.17	1.4699E-01
UTH-233	5548.	245.34	4.6794E+01
PA-234M	50.	1001.03	2.5166E+00
TH-234	4173.	63.29	3.0583E+00
U-235	6661.	185.71	7.7969E-02
NP-237	4554.	86.48	3.9051E-01
U-237	4225.	59.54	4.7982E-01
NP-238	46.	984.45	5.6728E-02
NP-239	4626.	106.12	1.6515E-01
PU-239	5152.	129.30	5.4945E+02
AM-241	4225.	59.54	4.5860E-01
AM-243	4279.	74.67	1.0847E-01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 01:22:35.50 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27142
 Sample ID: S98T002921-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3919.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:24.53 sec
 Dead Time: 0.3%

Verified by:

11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-NOV-1998 22:51:42.02
 Decayed to: 23-NOV-1998 22:51:42.02
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.68*	181351	1.59	1323.54	1315	17	0.5		CS-137	68.3

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002921-DUP

Acquisition date : 23-NOV-1998 22:51:42

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	6.832E+01	6.833E+01	0.033E+01	0.48	
Total Activity :			6.832E+01	6.833E+01			

Grand Total Activity :	6.832E+01	6.833E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002921-DUP

Acquisition date : 23-NOV-1998 22:51:42

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	5874.	477.59	8.4839E-01
NA-24	20.	1368.55	1.4967E-02
K-40	383.	1460.75	5.5003E-01
AR-41	29.	1293.60	2.4522E-02
SC-46	67.	1120.55	2.0554E-02
CR-51	4787.	320.08	5.7709E-01
MN-54	68.	834.83	1.6111E-02
CO-56	57.	846.76	1.5045E-02
CO-58	61.	810.78	1.4982E-02
FE-59	39.	1099.25	2.8032E-02
CO-60	29.	1332.50	1.6300E-02
ZN-65	36.	1115.55	3.0531E-02
SE-75	5306.	264.66	9.0920E-02
KR-85	2967.	514.00	1.5287E+01
SR-85	2966.	514.01	6.9133E-02
Y-88	5.	1836.06	9.8231E-03
Y-91	32.	1204.67	5.2025E+00
NB-94	64.	871.09	1.6260E-02
NB-95	64.	765.78	1.4583E-02
ZR-95	50.	756.73	2.3504E-02
ZRNB-95	78.	724.18	6.9795E-02
RU-103	3602.	497.08	7.8181E-02
RURH-106	1162.	621.93	9.9050E-01
AG-108m	76.	722.94	1.6478E-02
CD-109	4458.	88.03	1.2865E+00
AG-110M	4803.	657.76	1.0975E-01
SN-113	5084.	391.69	1.0896E-01
TE-123m	5306.	159.00	4.3582E-02
SB-124	1178.	602.73	4.9259E-02
SB-125	5915.	427.89	2.7416E-01
TE-125m	4798.	109.27	1.3119E+01
I-129	4394.	39.60	0.0000E+00
I-131	4783.	364.48	7.9259E-02
XE-131m	5139.	163.93	1.8693E+00
BA-133	4818.	356.02	1.0183E-01
CS-134	1166.	604.70	4.9223E-02
CS-136	63.	818.51	1.5303E-02
CS-138	18.	1435.86	6.0499E-02
CE-139	5139.	165.85	4.6004E-02
BA-140	2131.	537.31	2.4032E-01
LA-140	12.	1596.21	1.3702E-02
CE-141	5213.	145.44	7.3390E-02
CE-144	5237.	133.51	3.1855E-01
CEPR-144	5237.	133.51	6.3653E-01
EU-152	5186.	121.78	1.2679E-01
EU-154	27.	1274.51	4.3790E-02
EU-155	4520.	86.54	1.4689E-01
HF-181	4651.	482.18	9.7911E-02
TA-182	4164.	67.75	2.1189E-01
HG-203	5065.	279.20	6.6278E-02
BI-207	1262.	569.70	4.8523E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002921-DUP

Acquisition date : 23-NOV-1998 22:51:42

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	5209.	277.36	8.6294E-01
PB-210	4520.	46.50	0.0000E+00
BI-212	76.	727.18	2.2680E-01
PB-212	5947.	238.63	1.1964E-01
BI-214	1174.	609.31	1.0828E-01
PB-214	4790.	351.92	6.9285E-01
RA-224	5940.	240.99	1.3363E+00
RA-226	6917.	186.10	1.3071E+00
AC-228	82.	911.21	7.1254E-02
TH-228	4440.	84.37	3.9544E+00
TH-229	4481.	88.47	1.8693E-01
U-232	4342.	57.78	1.0372E+02
PA-233	4851.	312.17	1.4880E-01
UTH-233	5740.	245.34	4.7587E+01
PA-234M	51.	1001.03	2.5332E+00
TH-234	4203.	63.29	3.0695E+00
U-235	6808.	185.71	7.8822E-02
NP-237	4521.	86.48	3.8911E-01
U-237	4288.	59.54	4.8331E-01
NP-238	40.	984.45	5.3201E-02
NP-239	4646.	106.12	1.6551E-01
PU-239	5232.	129.30	5.5364E+02
AM-241	4288.	59.54	4.6193E-01
AM-243	4350.	74.67	1.0936E-01

*****

* 222-S Laboratory Counting Room 24-NOV-1998 09:19:12.29 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27142
 Sample ID: SS98TT2922-SAM
 Sample Size: 1.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3922.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.24 sec
 Dead Time: 0.1%

Verified by:

11-24-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 24-NOV-1998 06:35:20.62
 Decayed to: 24-NOV-1998 06:35:20.62
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
4	25.52	714	1.11	51.79	48	34	14.3	4.63E+01		
4	26.96	657	1.34	54.67	48	34	19.4			
4	33.13	299	1.02	67.00	48	34	34.0			
0	474.21	169	2.15	948.70	946	9	78.9			
0	661.73*	36677	1.57	1323.64	1316	16	1.1		CS-137	138.
0	1764.01*	18	2.24	3528.64	3519	18118.0				

Summary of Nuclide Activity
Sample ID : SS98TT2922-SAM

HNF-1663 REV.0

Page : 2

Acquisition date : 24-NOV-1998 06:35:20

Total number of lines in spectrum	6	
Number of unidentified lines	4	
Number of lines tentatively identified by NID	2	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma Error	2-Sigma
			Uncorrected uCi/L	Decay Corr uCi/L			
CS-137	30.00Y	1.00	1.382E+02	1.382E+02	0.015E+02		1.08
Total Activity :			1.382E+02	1.382E+02			

Grand Total Activity : 1.382E+02 1.382E+02

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : SS98TT2922-SAM

Acquisition date : 24-NOV-1998 06:35:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1317.	477.59	4.0511E+00
NA-24	18.	1368.55	1.4103E-01
K-40	393.	1460.75	5.5685E+00
AR-41	28.	1293.60	2.4136E-01
SC-46	67.	1120.55	2.0643E-01
CR-51	1125.	320.08	2.8229E+00
MN-54	54.	834.83	1.4397E-01
CO-56	47.	846.76	1.3740E-01
CO-58	52.	810.78	1.3961E-01
FE-59	42.	1099.25	2.8817E-01
CO-60	30.	1332.50	1.6520E-01
ZN-65	33.	1115.55	2.9149E-01
SE-75	1318.	264.66	4.5677E-01
KR-85	749.	514.00	7.7607E+01
SR-85	749.	514.01	3.5103E-01
Y-88	9.	1836.06	1.2504E-01
Y-91	39.	1204.67	5.7100E+01
NB-94	58.	871.09	1.5514E-01
NB-95	54.	765.78	1.3479E-01
ZR-95	66.	756.73	2.6837E-01
ZRNB-95	69.	724.18	6.5869E-01
RU-103	858.	497.08	3.8540E-01
RURH-106	258.	621.93	4.7576E+00
AG-108m	78.	722.94	1.6654E-01
CD-109	1664.	88.03	7.9026E+00
AG-110M	974.	657.76	4.9912E-01
SN-113	1174.	391.69	5.2818E-01
TE-123m	1599.	159.00	2.4081E-01
SB-124	284.	602.73	2.4596E-01
SB-125	1316.	427.89	1.3038E+00
TE-125m	1652.	109.27	7.7441E+01
I-129	1459.	39.60	0.0000E+00
I-131	1124.	364.48	3.8761E-01
XE-131m	1527.	163.93	1.0258E+01
BA-133	1100.	356.02	4.9091E-01
CS-134	283.	604.70	2.4666E-01
CS-136	62.	818.51	1.5277E-01
CS-138	19.	1435.86	6.1855E-01
CE-139	1534.	165.85	2.5305E-01
BA-140	514.	537.31	1.1950E+00
LA-140	11.	1596.21	1.2999E-01
CE-141	1645.	145.44	4.1483E-01
CE-144	1680.	133.51	1.8155E+00
CEPR-144	1681.	133.51	3.6280E+00
EU-152	1728.	121.78	7.3605E-01
EU-154	50.	1274.51	5.8134E-01
EU-155	1705.	86.54	9.0700E-01
HF-181	1078.	482.18	4.7566E-01
TA-182	1480.	67.75	1.2707E+00
HG-203	1233.	279.20	3.2969E-01
BI-207	375.	569.70	2.6809E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : SS98TT2922-SAM

Acquisition date : 24-NOV-1998 06:35:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1272.	277.36	4.2989E+00
PB-210	1448.	46.50	0.0000E+00
BI-212	64.	727.18	2.0945E+00
PB-212	1586.	238.63	6.2227E-01
BI-214	403.	609.31	6.4208E-01
PB-214	1125.	351.92	3.3813E+00
RA-224	1520.	240.99	6.8099E+00
RA-226	1797.	186.10	6.7069E+00
AC-228	89.	911.21	7.4005E-01
TH-228	1711.	84.37	2.4680E+01
TH-229	1642.	88.47	1.1380E+00
U-232	1518.	57.78	6.1701E+02
PA-233	1196.	312.17	7.4512E-01
UTH-233	1438.	245.34	2.3998E+02
PA-234M	48.	1001.03	2.4640E+01
TH-234	1428.	63.29	1.8005E+01
U-235	1769.	185.71	4.0452E-01
NP-237	1707.	86.48	2.4034E+00
U-237	1528.	59.54	2.9025E+00
NP-238	60.	984.45	6.3993E-01
NP-239	1630.	106.12	9.8610E-01
PU-239	1660.	129.30	3.1379E+03
AM-241	1528.	59.54	2.7740E+00
AM-243	1482.	74.67	6.4235E-01

*****
 * 222-S Laboratory Counting Room 24-NOV-1998 11:49:44.20 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27142
 Sample ID: S98T2922-DUP
 Sample Size: 1.00000E-04 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3923.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:06.33 sec
 Dead Time: 0.1%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 24-NOV-1998 09:19:04.79
 Decayed to: 24-NOV-1998 09:19:04.79
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	25.97	680	2.37	52.69	49	9	17.9			
0	661.74*	32182	1.62	1323.66	1316	16	1.1		CS-137	121.

Summary of Nuclide Activity
 Sample ID : S98T2922-DUP

Page : 2

Acquisition date : 24-NOV-1998 09:19:04

Total number of lines in spectrum	2	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	1	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.212E+02	1.212E+02	0.014E+02	1.15	
Total Activity :			1.212E+02	1.212E+02			

Grand Total Activity : 1.212E+02 1.212E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2922-DUP

Acquisition date : 24-NOV-1998 09:19:04

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1128.	477.59	3.7533E+00
NA-24	13.	1368.55	1.2352E-01
K-40	326.	1460.75	5.0861E+00
AR-41	20.	1293.60	2.1049E-01
SC-46	64.	1120.55	2.0163E-01
CR-51	967.	320.08	2.6200E+00
MN-54	58.	834.83	1.4994E-01
CO-56	67.	846.76	1.6224E-01
CO-58	50.	810.78	1.3628E-01
FE-59	33.	1099.25	2.5699E-01
CO-60	34.	1332.50	1.7387E-01
ZN-65	48.	1115.55	3.4716E-01
SE-75	1058.	264.66	4.0996E-01
KR-85	668.	514.00	7.3360E+01
SR-85	667.	514.01	3.3159E-01
Y-88	10.	1836.06	1.3178E-01
Y-91	33.	1204.67	5.2956E+01
NB-94	41.	871.09	1.3233E-01
NB-95	49.	765.78	1.2899E-01
ZR-95	63.	756.73	2.6240E-01
ZRNB-95	70.	724.18	6.6437E-01
RU-103	703.	497.08	3.4970E-01
RURH-106	265.	621.93	4.8156E+00
AG-108m	67.	722.94	1.5504E-01
CD-109	1241.	88.03	6.8400E+00
AG-110M	810.	657.76	4.5609E-01
SN-113	1000.	391.69	4.8813E-01
TE-123m	1301.	159.00	2.1759E-01
SB-124	257.	602.73	2.3465E-01
SB-125	1179.	427.89	1.2354E+00
TE-125m	1333.	109.27	6.9670E+01
I-129	1209.	39.60	0.0000E+00
I-131	942.	364.48	3.5534E-01
XE-131m	1248.	163.93	9.2891E+00
BA-133	948.	356.02	4.5630E-01
CS-134	278.	604.70	2.4441E-01
CS-136	59.	818.51	1.4930E-01
CS-138	15.	1435.86	5.5688E-01
CE-139	1255.	165.85	2.2918E-01
BA-140	433.	537.31	1.0990E+00
LA-140	16.	1596.21	1.5117E-01
CE-141	1286.	145.44	3.6743E-01
CE-144	1296.	133.51	1.5976E+00
CEPR-144	1296.	133.51	3.1919E+00
EU-152	1287.	121.78	6.3668E-01
EU-154	28.	1274.51	4.4875E-01
EU-155	1258.	86.54	7.8096E-01
HF-181	906.	482.18	4.3667E-01
TA-182	1165.	67.75	1.1296E+00
HG-203	1118.	279.20	3.1420E-01
BI-207	326.	569.70	2.5041E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2922-DUP

Acquisition date : 24-NOV-1998 09:19:04

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1089.	277.36	3.9827E+00
PB-210	1225.	46.50	0.0000E+00
BI-212	80.	727.18	2.3240E+00
PB-212	1371.	238.63	5.7912E-01
BI-214	340.	609.31	5.9142E-01
PB-214	1038.	351.92	3.2503E+00
RA-224	1274.	240.99	6.2437E+00
RA-226	1511.	186.10	6.1575E+00
AC-228	72.	911.21	6.6764E-01
TH-228	1239.	84.37	2.1047E+01
TH-229	1218.	88.47	9.8246E-01
U-232	1244.	57.78	5.5933E+02
PA-233	957.	312.17	6.6793E-01
UTH-233	1277.	245.34	2.2636E+02
PA-234M	43.	1001.03	2.3361E+01
TH-234	1060.	63.29	1.5548E+01
U-235	1527.	185.71	3.7614E-01
NP-237	1258.	86.48	2.0685E+00
U-237	1229.	59.54	2.6073E+00
NP-238	42.	984.45	5.4505E-01
NP-239	1272.	106.12	8.7267E-01
PU-239	1379.	129.30	2.8642E+03
AM-241	1229.	59.54	2.4919E+00
AM-243	1147.	74.67	5.6625E-01

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 15 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 4486 THRU 4662)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	GAMMA ENERGY ANALYSIS, GEA WORKLIST 27143 THRU 28056		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSE Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955, 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17Sec 5
of 17

TABLE OF CONTENTS (Continued)

Sec 5
of 17
Cont.

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

Sec 7
of 17
Cont.

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 27143

Analyst: crj

Instrument: GEA01

Book#: 28B57

Method: LA-548-121 Rev/Mod _____

Worklist Comment: BY105, @GEA-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit	
1 STD	0	@GEA-01	CO60-02	SOLID	4.519e-03	4.78e-03	105.776	% Recovery	
1 STD	0	@GEA-01	CO60-02E	SOLID	1	3.16	3.160	% Ct Error	
1 STD	0	@GEA-01	CS13702	SOLID	7.450e-03	7.89e-03	105.905	% Recovery	
1 STD	0	@GEA-01	CS13702E	SOLID	1	2.78	2.780	% Ct Error	
2 BLNK-PREP	0	@GEA-01	CO60-02	SOLID	1	<1.35e-2		uCi/g	
2 BLNK-PREP	0	@GEA-01	CS13702	SOLID	1	4.27e-02	0.043	uCi/g	
4 DUP	S98T002986	0 F	@GEA-01	CO60-02	SOLID	<1.36e-2	<1.38e-2		RPD
4 DUP	S98T002986	0 F	@GEA-01	CO60-02E	SOLID	1	n/a		% Ct Error
4 DUP	S98T002986	0 F	@GEA-01	CS13702	SOLID	7.56e+01	4.85e+01	43.674	RPD
4 DUP	S98T002986	0 F	@GEA-01	CS13702E	SOLID	1.00	0.420	0.420	% Ct Error
5 SAMPLE	S98T002987	0 F	@GEA-01	CO60-02	SOLID	N/A	<1.469e-02	1.47e-002	uCi/g
5 SAMPLE	S98T002987	0 F	@GEA-01	CO60-02E	SOLID	N/A	n/a		% Ct. Error
5 SAMPLE	S98T002987	0 F	@GEA-01	CS13702	SOLID	N/A	6.561e+01		uCi/g
5 SAMPLE	S98T002987	0 F	@GEA-01	CS13702E	SOLID	N/A	0.360		% Ct. Error
6 DUP	S98T002987	0 F	@GEA-01	CO60-02	SOLID	<1.47e-2	<1.46e-2		RPD
6 DUP	S98T002987	0 F	@GEA-01	CO60-02E	SOLID	1	n/a		% Ct Error
6 DUP	S98T002987	0 F	@GEA-01	CS13702	SOLID	6.56e+01	6.25e+01	4.840	RPD
6 DUP	S98T002987	0 F	@GEA-01	CS13702E	SOLID	1.00	0.370	0.370	% Ct Error

Comments Section:

Comments for sample# S98T002987 and test @GEA-01 .
DL=0 => n/a.

Final page for worklist# 27143

Analyst Signature	Date	Analyst Signature	Date
	20 Nov 98		
Reviewer Signature	Date		

Sample S98T002986 Rejected for
High RPD. Resun worklist is # 27278

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27143

Analyst: CRJ Instrument: GEA00 + Book# 28857Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105, @GEA-01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	SOLID		
2 BLNK-PREP			@GEA-01	SOLID		
3 SAMPLE	S98T002986 0 F	@GEA-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
4 DUP	S98T002986 0 F	@GEA-01	SOLID			
5 SAMPLE	S98T002987 0 F	@GEA-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
6 DUP	S98T002987 0 F	@GEA-01	SOLID			

Final page for worklist # 27143

CRJ 11/19/98
Analyst Signature DateEm Bucher 11-20-98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 19-NOV-1998 16:00:19.47 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: WL27143-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Neu Jight

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1101.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:14.88 sec
 Dead Time: 0.5%

Verified by:

John Relyea 20 Nov 98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 15:09:34.97
 Decayed to: 19-NOV-1998 15:09:34.97
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	31.98	1046	1.05	64.07	60	8 11.7			
0	36.24	272	0.90	72.61	69	9 41.3			
0	59.35	379	1.01	118.81	114	9 30.3		AM-241	0.263
0	238.51*	82	0.98	477.11	474	7102.4			
0	661.73*	6517	1.39	1323.56	1316	16 2.8		CS-137	7.89
0	1173.16	2764	1.86	2346.62	2337	18 4.6		CO-60	4.76
0	1332.46	2521	1.91	2665.32	2656	20 4.4		CO-60	4.81

Summary of Nuclide Activity

Page : 2

Sample ID : WL27143-STD

Acquisition date : 19-NOV-1998 15:09:34

Total number of lines in spectrum	7	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	5	71.43%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected	Decay Corr			
			uCi/L	uCi/L	2-Sigma Error	%Error	
CO-60	5.27Y	1.00	4.784E+00	4.784E+00	0.151E+00	3.16	
CS-137	30.00Y	1.00	7.891E+00	7.891E+00	0.219E+00	2.78	
AM-241	432.20Y*****		2.628E-01	2.628E-01	0.797E-01	30.32	
Total Activity :			1.294E+01	1.294E+01			

Grand Total Activity :	1.294E+01	1.294E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27143-STD

Acquisition date : 19-NOV-1998 15:09:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	418.	477.59	7.1977E-01
NA-24	14.	1368.55	4.0405E-02
K-40	588.	1460.75	2.2169E+00
AR-41	36.	1293.60	6.7136E-02
SC-46	173.	1120.55	1.0578E-01
CR-51	434.	320.08	5.2616E-01
MN-54	213.	834.83	9.0228E-02
CO-56	176.	846.76	8.3457E-02
CO-58	180.	810.78	8.1362E-02
FE-59	163.	1099.25	1.7867E-01
ZN-65	168.	1115.55	2.0473E-01
SE-75	519.	264.66	8.4439E-02
KR-85	329.	514.00	1.6359E+01
SR-85	329.	514.01	7.3918E-02
Y-88	9.	1836.06	4.0032E-02
Y-91	77.	1204.67	2.5516E+01
NB-94	176.	871.09	8.5466E-02
NB-95	179.	765.78	7.6885E-02
ZR-95	213.	756.73	1.5110E-01
ZRNB-95	200.	724.18	3.5213E-01
RU-103	331.	497.08	7.5913E-02
RURH-106	236.	621.93	1.4553E+00
AG-108m	184.	722.94	8.1159E-02
CD-109	626.	88.03	8.2957E-01
AG-110M	304.	657.76	9.0903E-02
SN-113	413.	391.69	9.6563E-02
TE-123m	728.	159.00	4.8569E-02
SB-124	210.	602.73	6.7724E-02
SB-125	392.	427.89	2.2200E-01
TE-125m	650.	109.27	1.1472E+01
I-129	461.	39.60	3.9225E-01
I-131	436.	364.48	7.3369E-02
XE-131m	690.	163.93	2.0671E+00
BA-133	398.	356.02	8.9870E-02
CS-134	208.	604.70	6.7772E-02
CS-136	179.	818.51	8.1842E-02
CS-138	19.	1435.86	1.0001E-01
CE-139	687.	165.85	5.0902E-02
BA-140	222.	537.31	2.5135E-01
LA-140	9.	1596.21	3.9343E-02
CE-141	735.	145.44	8.0935E-02
CE-144	751.	133.51	3.4047E-01
CEPR-144	751.	133.51	6.8029E-01
EU-152	698.	121.78	1.2302E-01
EU-154	38.	1274.51	1.6836E-01
EU-155	606.	86.54	8.9615E-02
HF-181	359.	482.18	8.6757E-02
TA-182	532.	67.75	6.5672E-02
HG-203	512.	279.20	6.2925E-02
BI-207	225.	569.70	6.6465E-02
TL-208	487.	277.36	7.8851E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27143-STD

Acquisition date : 19-NOV-1998 15:09:34

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	485.	46.50	6.9061E-01
BI-212	218.	727.18	1.2093E+00
PB-212	728.	238.63	1.2328E-01
BI-214	328.	609.31	1.8545E-01
PB-214	528.	351.92	3.1618E-01
RA-224	595.	240.99	1.2404E+00
RA-226	764.	186.10	1.3160E+00
AC-228	331.	911.21	4.5375E-01
TH-228	649.	84.37	2.3896E+00
TH-229	629.	88.47	1.2163E-01
U-232	627.	57.78	1.4896E+01
PA-233	480.	312.17	1.4124E-01
UTH-233	551.	245.34	4.3575E+01
PA-234M	199.	1001.03	1.5797E+01
TH-234	569.	63.29	7.3858E-01
U-235	769.	185.71	8.0246E-02
NP-237	608.	86.48	2.3735E-01
U-237	883.	59.54	1.0180E-01
NP-238	175.	984.45	3.4387E-01
NP-239	660.	106.12	1.4162E-01
PU-239	740.	129.30	5.7572E+02
AM-243	634.	74.67	4.9213E-02

*****

* 222-S Laboratory Counting Room 19-NOV-1998 18:41:04.95 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: WL27143-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

BL Under

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]lg1102.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:32.70 sec
 Dead Time: 0.4%

Verified by:

John Relyea 20 Nov 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 16:10:01.97
 Decayed to: 19-NOV-1998 16:10:01.97
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.83	133	0.80	25.78	24	6 78.0			
0	16.01	227	1.26	32.14	29	7 51.7			
0	661.84*	209	1.50	1323.78	1318	12 37.9			
0	1409.51	19	1.28	2819.47	2813	9 129.4			
0	1730.68	21	1.62	3462.08	3454	12 98.0			
0	1764.64*	19	1.97	3530.04	3523	13154.6			
								CS-137	8.428E-02

Summary of Nuclide Activity

Page : 2

Sample ID : WL27143-BLK

Acquisition date : 19-NOV-1998 16:10:01

Total number of lines in spectrum	6	
Number of unidentified lines	3	
Number of lines tentatively identified by NID	3	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma Error	2-Sigma	Flags
			Uncorrected	Decay Corr				
			uCi/L	uCi/L			%Error	
CS-137	30.00Y	1.00	8.428E-02	8.428E-02		3.192E-02	37.88	
Total Activity :			8.428E-02	8.428E-02				

Grand Total Activity :	8.428E-02	8.428E-02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27143-BLK

Acquisition date : 19-NOV-1998 16:10:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	337.	477.59	2.1632E-01
NA-24	64.	1368.55	2.7343E-02
K-40	1777.	1460.75	1.2714E+00
AR-41	88.	1293.60	4.5011E-02
SC-46	248.	1120.55	4.1893E-02
CR-51	623.	320.08	2.0950E-01
MN-54	172.	834.83	2.7100E-02
CO-56	155.	846.76	2.6129E-02
CO-58	168.	810.78	2.6261E-02
FE-59	139.	1099.25	5.5246E-02
CO-60	71.	1332.50	2.6660E-02
ZN-65	129.	1115.55	6.0254E-02
SE-75	825.	264.66	3.5331E-02
KR-85	394.	514.00	5.9529E+00
SR-85	393.	514.01	2.6889E-02
Y-88	20.	1836.06	1.9003E-02
Y-91	118.	1204.67	1.0400E+01
NB-94	141.	871.09	2.5638E-02
NB-95	206.	765.78	2.7395E-02
ZR-95	193.	756.73	4.8100E-02
ZRNB-95	199.	724.18	1.1719E-01
RU-103	291.	497.08	2.3769E-02
RURH-106	201.	621.93	4.4894E-01
AG-108m	187.	722.94	2.7253E-02
CD-109	1187.	88.03	3.7840E-01
AG-110M	216.	657.76	2.5685E-02
SN-113	456.	391.69	3.3795E-02
TE-123m	1351.	159.00	2.1941E-02
SB-124	212.	602.73	2.2729E-02
SB-125	390.	427.89	7.3806E-02
TE-125m	1329.	109.27	5.4314E+00
I-129	736.	39.60	1.6428E-01
I-131	479.	364.48	2.5692E-02
XE-131m	1304.	163.93	9.4350E-01
BA-133	486.	356.02	3.3024E-02
CS-134	235.	604.70	2.3993E-02
CS-136	149.	818.51	2.4998E-02
CS-138	52.	1435.86	1.0773E-01
CE-139	1280.	165.85	2.3029E-02
BA-140	278.	537.31	9.3393E-02
LA-140	33.	1596.21	2.3111E-02
CE-141	1411.	145.44	3.7206E-02
CE-144	1486.	133.51	1.5865E-01
CEPR-144	1485.	133.51	3.1699E-01
EU-152	1381.	121.78	5.7325E-02
EU-154	101.	1274.51	8.8339E-02
EU-155	1238.	86.54	4.2380E-02
HF-181	349.	482.18	2.8550E-02
TA-182	1007.	67.75	2.9926E-02
HG-203	806.	279.20	2.6199E-02
BI-207	250.	569.70	2.3304E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27143-BLK

Acquisition date : 19-NOV-1998 16:10:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	807.	277.36	3.3640E-01
PB-210	950.	46.50	3.2084E-01
BI-212	271.	727.18	4.4721E-01
PB-212	1248.	238.63	5.3532E-02
BI-214	588.	609.31	8.2086E-02
PB-214	846.	351.92	2.9519E-01
RA-224	1036.	240.99	5.4612E-01
RA-226	1344.	186.10	5.7893E-01
AC-228	337.	911.21	1.5256E-01
TH-228	1318.	84.37	1.1278E+00
TH-229	1173.	88.47	5.5049E-02
U-232	1086.	57.78	6.4990E+00
PA-233	668.	312.17	5.5390E-02
UTH-233	843.	245.34	1.7878E+01
PA-234M	142.	1001.03	4.4714E+00
TH-234	1121.	63.29	3.4324E-01
U-235	1334.	185.71	3.5065E-02
NP-237	1244.	86.48	1.1236E-01
U-237	1123.	59.54	3.8321E-02
NP-238	130.	984.45	1.0056E-01
NP-239	1324.	106.12	6.7137E-02
PU-239	1417.	129.30	2.6405E+02
AM-241	1123.	59.54	3.6629E-02
AM-243	1291.	74.67	2.3245E-02

*****
 * 222-S Laboratory Counting Room 19-NOV-1998 23:48:26.49 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: S98T2986-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SL Uddel

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1104.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:32:04.49 sec
 Dead Time: 1.4%

Verified by:

John Relyea 20 Nov 98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 19-NOV-1998 21:15:53.12
 Decayed to: 19-NOV-1998 21:15:53.12
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	21.88	1979	1.07	43.88	40	8	19.7			
0	31.94	62785	1.02	64.00	60	10	1.3			
0	36.48	17579	1.05	73.08	69	9	3.0			
0	661.68*	368993	1.44	1323.46	1315	16	0.3		CS-137	149.
0	785.83	88	1.96	1571.80	1567	9	46.2			
0	1382.51	105	9.73	2765.45	2745	42	80.7			
0	1523.35	15	1.33	3047.23	3044	6	63.2			
0	1615.67	20	1.10	3231.95	3225	13	86.7			

DATA
NOT USED

Summary of Nuclide Activity
Sample ID : S98T2986-SAM

HNF-1663 REV. 0

Page : 2

Acquisition date : 19-NOV-1998 21:15:53

Total number of lines in spectrum	8	
Number of unidentified lines	6	
Number of lines tentatively identified by NID	2	25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.489E+02	1.489E+02	0.005E+02	0.34	
Total Activity :			1.489E+02	1.489E+02			

Grand Total Activity : 1.489E+02 1.489E+02

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2986-SAM

Acquisition date : 19-NOV-1998 21:15:53

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	11797.	477.59	1.2469E+00
NA-24	64.	1368.55	2.7508E-02
K-40	1687.	1460.75	1.2392E+00
AR-41	90.	1293.60	4.5524E-02
SC-46	228.	1120.55	4.0256E-02
CR-51	9460.	320.08	8.0231E-01
MN-54	186.	834.83	2.8183E-02
CO-56	154.	846.76	2.6090E-02
CO-58	153.	810.78	2.5119E-02
FE-59	115.	1099.25	5.0436E-02
CO-60	72.	1332.50	2.6746E-02
ZN-65	128.	1115.55	5.9947E-02
SE-75	9598.	264.66	1.1879E-01
KR-85	6009.	514.00	2.2750E+01
SR-85	6007.	514.01	1.0289E-01
Y-88	23.	1836.06	2.0319E-02
Y-91	136.	1204.67	1.1120E+01
NB-94	144.	871.09	2.5893E-02
NB-95	183.	765.78	2.5901E-02
ZR-95	205.	756.73	4.9564E-02
ZRNB-95	223.	724.18	1.2368E-01
RU-103	7337.	497.08	1.1625E-01
RURH-106	1953.	621.93	1.3628E+00
AG-108m	212.	722.94	2.8953E-02
CD-109	8204.	88.03	9.8468E-01
AG-110M	10010.	657.76	1.6920E-01
SN-113	10387.	391.69	1.5791E-01
TE-123m	9442.	159.00	5.7437E-02
SB-124	2035.	602.73	6.8538E-02
SB-125	12184.	427.89	4.0303E-01
TE-125m	8776.	109.27	1.3825E+01
I-129	7127.	39.60	5.0399E-01
I-131	9831.	364.48	1.1405E-01
XE-131m	9355.	163.93	2.5017E+00
BA-133	9623.	356.02	1.4399E-01
CS-134	2068.	604.70	6.9397E-02
CS-136	183.	818.51	2.7622E-02
CS-138	51.	1435.86	1.0758E-01
CE-139	9377.	165.85	6.1708E-02
BA-140	4224.	537.31	3.5519E-01
LA-140	41.	1596.21	2.5336E-02
CE-141	9415.	145.44	9.5197E-02
CE-144	9319.	133.51	3.9379E-01
CEPR-144	9319.	133.51	7.8684E-01
EU-152	9497.	121.78	1.4888E-01
EU-154	96.	1274.51	8.6267E-02
EU-155	8141.	86.54	1.0761E-01
HF-181	9137.	482.18	1.4247E-01
TA-182	7602.	67.75	8.1281E-02
HG-203	9403.	279.20	8.8223E-02
BI-207	2564.	569.70	7.2861E-02

DATA
NOT USED

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2986-SAM

Acquisition date : 19-NOV-1998 21:15:53

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	9422.	277.36	1.1331E+00
PB-210	7211.	46.50	8.7347E-01
BI-212	271.	727.18	4.4712E-01
PB-212	10303.	238.63	1.5220E-01
BI-214	2415.	609.31	1.6438E-01
PB-214	9980.	351.92	1.0092E+00
RA-224	10078.	240.99	1.6826E+00
RA-226	13494.	186.10	1.8150E+00
AC-228	342.	911.21	1.5363E-01
TH-228	8341.	84.37	2.8101E+00
TH-229	8262.	88.47	1.4455E-01
U-232	7389.	57.78	1.6768E+01
PA-233	9226.	312.17	2.0256E-01
UTH-233	9954.	245.34	6.0578E+01
PA-234M	144.	1001.03	4.4977E+00
TH-234	7605.	63.29	8.8459E-01
U-235	13322.	185.71	1.0962E-01
NP-237	8156.	86.48	2.8488E-01
U-237	7481.	59.54	9.7865E-02
NP-238	148.	984.45	1.0683E-01
NP-239	8346.	106.12	1.6697E-01
PU-239	9359.	129.30	6.7241E+02
AM-241	7481.	59.54	9.3540E-02
AM-243	8161.	74.67	5.7882E-02

DATA
NOT USED

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 03:04:16.80 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: S98T2986-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

gllali

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1106.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:31.29 sec
 Dead Time: 1.0%

Verified by:

John Relyea 20 NOV 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 00:32:18.20
 Decayed to: 20-NOV-1998 00:32:18.20
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	21.85	1327	1.12	43.82	40	8 24.3			
0	31.95	41158	1.02	64.02	60	9 1.5			
0	36.48	11627	1.01	73.07	69	9 3.7			
0	661.69*	238646	1.43	1323.47	1315	16 0.4		CS-137	96.3
0	1409.34	44	3.75	2819.13	2808	19 91.6			
0	1661.30	27	0.93	3323.25	3317	12 74.1			

DATA
 NOT USED

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2986-DUP

Acquisition date : 20-NOV-1998 00:32:18

Total number of lines in spectrum	6	
Number of unidentified lines	4	
Number of lines tentatively identified by NID	2	33.33%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	9.631E+01	9.631E+01	0.041E+01	0.42	
Total Activity :			9.631E+01	9.631E+01			

Grand Total Activity :	9.631E+01	9.631E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2986-DUP

Acquisition date : 20-NOV-1998 00:32:18

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	7707.	477.59	1.0091E+00
NA-24	62.	1368.55	2.7010E-02
K-40	1773.	1460.75	1.2702E+00
AR-41	98.	1293.60	4.7303E-02
SC-46	252.	1120.55	4.2206E-02
CR-51	6292.	320.08	6.5522E-01
MN-54	187.	834.83	2.8207E-02
CO-56	145.	846.76	2.5349E-02
CO-58	174.	810.78	2.6695E-02
FE-59	149.	1099.25	5.7211E-02
CO-60	75.	1332.50	2.7331E-02
ZN-65	127.	1115.55	5.9722E-02
SE-75	6582.	264.66	9.8491E-02
KR-85	4127.	514.00	1.8883E+01
SR-85	4126.	514.01	8.5397E-02
Y-88	20.	1836.06	1.9003E-02
Y-91	151.	1204.67	1.1678E+01
NB-94	133.	871.09	2.4983E-02
NB-95	217.	765.78	2.8117E-02
ZR-95	196.	756.73	4.8515E-02
ZRNB-95	217.	724.18	1.2228E-01
RU-103	4935.	497.08	9.5486E-02
RURH-106	1305.	621.93	1.1173E+00
AG-108m	205.	722.94	2.8475E-02
CD-109	5799.	88.03	8.2887E-01
AG-110M	6520.	657.76	1.3674E-01
SN-113	6897.	391.69	1.2884E-01
TE-123m	6402.	159.00	4.7355E-02
SB-124	1389.	602.73	5.6783E-02
SB-125	7925.	427.89	3.2544E-01
TE-125m	6116.	109.27	1.1555E+01
I-129	4839.	39.60	4.1588E-01
I-131	6572.	364.48	9.3370E-02
XE-131m	6663.	163.93	2.1135E+00
BA-133	6458.	356.02	1.1811E-01
CS-134	1342.	604.70	5.6070E-02
CS-136	163.	818.51	2.6133E-02
CS-138	44.	1435.86	1.0035E-01
CE-139	6690.	165.85	5.2181E-02
BA-140	2833.	537.31	2.9148E-01
LA-140	51.	1596.21	2.7958E-02
CE-141	6626.	145.44	7.9951E-02
CE-144	6527.	133.51	3.2994E-01
CEPR-144	6527.	133.51	6.5928E-01
EU-152	6598.	121.78	1.2424E-01
EU-154	90.	1274.51	8.3496E-02
EU-155	5855.	86.54	9.1365E-02
HF-181	6142.	482.18	1.1696E-01
TA-182	5182.	67.75	6.7200E-02
HG-203	6462.	279.20	7.3224E-02
BI-207	1764.	569.70	6.0566E-02

DATA
NOT USED

4502

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2986-DUP

Acquisition date : 20-NOV-1998 00:32:18

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	6477.	277.36	9.4063E-01
PB-210	4991.	46.50	7.2765E-01
BI-212	268.	727.18	4.4477E-01
PB-212	7197.	238.63	1.2735E-01
BI-214	1873.	609.31	1.4500E-01
PB-214	6756.	351.92	8.2858E-01
RA-224	6897.	240.99	1.3935E+00
RA-226	9177.	186.10	1.4984E+00
AC-228	326.	911.21	1.5001E-01
TH-228	5930.	84.37	2.3722E+00
TH-229	5778.	88.47	1.2103E-01
U-232	5249.	57.78	1.4151E+01
PA-233	6392.	312.17	1.6881E-01
UTH-233	6702.	245.34	4.9772E+01
PA-234M	163.	1001.03	4.7785E+00
TH-234	5308.	63.29	7.4000E-01
U-235	9098.	185.71	9.0691E-02
NP-237	5868.	86.48	2.4192E-01
U-237	5199.	59.54	8.1690E-02
NP-238	126.	984.45	9.8937E-02
NP-239	6056.	106.12	1.4237E-01
PU-239	6686.	129.30	5.6897E+02
AM-241	5199.	59.54	7.8082E-02
AM-243	5912.	74.67	4.9320E-02

DATA
NOT USED

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 05:39:01.52 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: S98T2987-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Im Burela

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1107.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:55.71 sec
 Dead Time: 1.3%

Verified by:

John Rehyer 20 NOV 98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 03:06:36.13
 Decayed to: 20-NOV-1998 03:06:36.13
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	12.44	527	0.84	25.00	23	6	51.4			
0	21.81	1860	1.12	43.74	40	8	20.0			
0	31.95	56632	1.01	64.02	60	9	1.3			
0	36.44	16043	1.03	73.00	69	9	3.1			
0	661.68*	330400	1.43	1323.45	1315	16	0.4		CS-137	133.

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2987-SAM

Acquisition date : 20-NOV-1998 03:06:36

Total number of lines in spectrum	5	
Number of unidentified lines	4	
Number of lines tentatively identified by NID	1	20.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.333E+02	1.333E+02	0.005E+02	0.36	
Total Activity :			1.333E+02	1.333E+02			

Grand Total Activity :	1.333E+02	1.333E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2987-SAM

Acquisition date : 20-NOV-1998 03:06:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	10673.	477.59	1.1864E+00
NA-24	61.	1368.55	2.6788E-02
K-40	1753.	1460.75	1.2632E+00
AR-41	102.	1293.60	4.8410E-02
SC-46	257.	1120.55	4.2623E-02
CR-51	8498.	320.08	7.6067E-01
MN-54	160.	834.83	2.6199E-02
CO-56	160.	846.76	2.6599E-02
CO-58	189.	810.78	2.7765E-02
FE-59	178.	1099.25	6.2150E-02
CO-60	90.	1332.50	2.9784E-02
ZN-65	123.	1115.55	5.8786E-02
SE-75	8890.	264.66	1.1436E-01
KR-85	5399.	514.00	2.1573E+01
SR-85	5397.	514.01	9.7568E-02
Y-88	23.	1836.06	2.0378E-02
Y-91	155.	1204.67	1.1855E+01
NB-94	142.	871.09	2.5704E-02
NB-95	193.	765.78	2.6564E-02
ZR-95	184.	756.73	4.7007E-02
ZRNB-95	220.	724.18	1.2305E-01
RU-103	6435.	497.08	1.0892E-01
RURH-106	1807.	621.93	1.3116E+00
AG-108m	218.	722.94	2.9314E-02
CD-109	7424.	88.03	9.3701E-01
AG-110M	9095.	657.76	1.6133E-01
SN-113	9573.	391.69	1.5164E-01
TE-123m	8558.	159.00	5.4698E-02
SB-124	1918.	602.73	6.6572E-02
SB-125	11110.	427.89	3.8494E-01
TE-125m	7942.	109.27	1.3155E+01
I-129	6501.	39.60	4.8148E-01
I-131	9048.	364.48	1.0944E-01
XE-131m	8492.	163.93	2.3842E+00
BA-133	8705.	356.02	1.3699E-01
CS-134	1885.	604.70	6.6304E-02
CS-136	158.	818.51	2.5736E-02
CS-138	44.	1435.86	1.0098E-01
CE-139	8665.	165.85	5.9333E-02
BA-140	3940.	537.31	3.4315E-01
LA-140	42.	1596.21	2.5642E-02
CE-141	8363.	145.44	8.9753E-02
CE-144	8536.	133.51	3.7699E-01
CEPR-144	8537.	133.51	7.5332E-01
EU-152	8584.	121.78	1.4159E-01
EU-154	108.	1274.51	9.0871E-02
EU-155	7651.	86.54	1.0435E-01
HF-181	8383.	482.18	1.3650E-01
TA-182	7014.	67.75	7.8092E-02
HG-203	8663.	279.20	8.4703E-02
BI-207	2273.	569.70	6.8640E-02

4506

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2987-SAM

Acquisition date : 20-NOV-1998 03:06:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	8579.	277.36	1.0815E+00
PB-210	6686.	46.50	8.4132E-01
BI-212	258.	727.18	4.3673E-01
PB-212	9443.	238.63	1.4575E-01
BI-214	2298.	609.31	1.6041E-01
PB-214	9230.	351.92	9.6986E-01
RA-224	9318.	240.99	1.6182E+00
RA-226	12326.	186.10	1.7351E+00
AC-228	377.	911.21	1.6113E-01
TH-228	7868.	84.37	2.7298E+00
TH-229	7359.	88.47	1.3647E-01
U-232	6933.	57.78	1.6246E+01
PA-233	8627.	312.17	1.9591E-01
UTH-233	9083.	245.34	5.7883E+01
PA-234M	138.	1001.03	4.4189E+00
TH-234	6958.	63.29	8.4640E-01
U-235	12166.	185.71	1.0478E-01
NP-237	7674.	86.48	2.7639E-01
U-237	6975.	59.54	9.4524E-02
NP-238	129.	984.45	1.0007E-01
NP-239	7659.	106.12	1.5999E-01
PU-239	8764.	129.30	6.5082E+02
AM-241	6975.	59.54	9.0347E-02
AM-243	7544.	74.67	5.5666E-02

*****
 * 222-S Laboratory Counting Room 20-NOV-1998 09:08:05.34 *
 *****

>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27143
 Sample ID: S98T2987-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Im Burch

>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1108.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:52.57 sec
 Dead Time: 1.2%

Verified by:

John Pelyea 20/NOV/98

>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 20-NOV-1998 06:35:43.00
 Decayed to: 20-NOV-1998 06:35:43.00
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	12.55	398	0.71	25.23	23	6	67.3			
3	20.66	348	0.62	41.44	40	8	60.4	1.70E+00		
3	22.03	1726	0.94	44.17	40	8	17.0			
0	31.95	54397	1.01	64.03	60	8	1.3			
0	36.45	15340	1.06	73.03	69	9	3.2			
0	661.67*	315338	1.44	1323.45	1315	16	0.4		CS-137	127.
0	1460.62*	89	1.86	2921.74	2912	20	115.2		K-40	0.572
0	1731.48	25	0.90	3463.70	3454	17	107.1			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2987-DUP

Acquisition date : 20-NOV-1998 06:35:43

Total number of lines in spectrum 8
 Number of unidentified lines 6
 Number of lines tentatively identified by NID 2 25.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~5.718E-01~~	~~5.718E-01~~	~~6.589E-01~~	~~115.23~~	~~879~~
CS-137	30.00Y	1.00	1.273E+02	1.273E+02	0.005E+02	0.37	11/20/98
Total Activity :			1.278E+02	1.278E+02			

Grand Total Activity : 1.278E+02 1.278E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2987-DUP

Acquisition date : 20-NOV-1998 06:35:43

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	10222.	477.59	1.1611E+00
NA-24	68.	1368.55	2.8246E-02
AR-41	90.	1293.60	4.5534E-02
SC-46	269.	1120.55	4.3548E-02
CR-51	8007.	320.08	7.3850E-01
MN-54	174.	834.83	2.7253E-02
CO-56	149.	846.76	2.5648E-02
CO-58	164.	810.78	2.5997E-02
FE-59	151.	1099.25	5.7458E-02
CO-60	90.	1332.50	2.9714E-02
ZN-65	159.	1115.55	6.6528E-02
SE-75	8437.	264.66	1.1142E-01
KR-85	5136.	514.00	2.1045E+01
SR-85	5134.	514.01	9.5180E-02
Y-88	28.	1836.06	2.2132E-02
Y-91	155.	1204.67	1.1832E+01
NB-94	140.	871.09	2.5542E-02
NB-95	192.	765.78	2.6487E-02
ZR-95	152.	756.73	4.2865E-02
ZRNB-95	232.	724.18	1.2612E-01
RU-103	6156.	497.08	1.0655E-01
RURH-106	1678.	621.93	1.2646E+00
AG-108m	224.	722.94	2.9723E-02
CD-109	7065.	88.03	9.1425E-01
AG-110M	8505.	657.76	1.5604E-01
SN-113	8898.	391.69	1.4622E-01
TE-123m	8246.	159.00	5.3699E-02
SB-124	1798.	602.73	6.4487E-02
SB-125	10621.	427.89	3.7642E-01
TE-125m	7705.	109.27	1.2959E+01
I-129	6328.	39.60	4.7509E-01
I-131	8299.	364.48	1.0485E-01
XE-131m	8219.	163.93	2.3458E+00
BA-133	8317.	356.02	1.3392E-01
CS-134	1795.	604.70	6.4726E-02
CS-136	170.	818.51	2.6667E-02
CS-138	58.	1435.86	1.1463E-01
CE-139	8321.	165.85	5.8151E-02
BA-140	3581.	537.31	3.2729E-01
LA-140	37.	1596.21	2.4200E-02
CE-141	8202.	145.44	8.8887E-02
CE-144	8229.	133.51	3.7018E-01
CEPR-144	8229.	133.51	7.3969E-01
EU-152	8283.	121.78	1.3910E-01
EU-154	94.	1274.51	8.5175E-02
EU-155	7250.	86.54	1.0159E-01
HF-181	8042.	482.18	1.3372E-01
TA-182	6678.	67.75	7.6214E-02
HG-203	8324.	279.20	8.3041E-02
BI-207	2175.	569.70	6.7160E-02
TL-208	8322.	277.36	1.0653E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2987-DUP

Acquisition date : 20-NOV-1998 06:35:43

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	6182.	46.50	8.0922E-01
BI-212	269.	727.18	4.4579E-01
PB-212	9001.	238.63	1.4232E-01
BI-214	2170.	609.31	1.5593E-01
PB-214	8702.	351.92	9.4159E-01
RA-224	8943.	240.99	1.5856E+00
RA-226	11571.	186.10	1.6814E+00
AC-228	369.	911.21	1.5931E-01
TH-228	7437.	84.37	2.6544E+00
TH-229	7047.	88.47	1.3357E-01
U-232	6581.	57.78	1.5831E+01
PA-233	8073.	312.17	1.8956E-01
UTH-233	8691.	245.34	5.6629E+01
PA-234M	150.	1001.03	4.5966E+00
TH-234	6769.	63.29	8.3488E-01
U-235	11469.	185.71	1.0176E-01
NP-237	7267.	86.48	2.6901E-01
U-237	6660.	59.54	9.2380E-02
NP-238	124.	984.45	9.8328E-02
NP-239	7601.	106.12	1.5939E-01
PU-239	8191.	129.30	6.2932E+02
AM-241	6660.	59.54	8.8297E-02
AM-243	7404.	74.67	5.5149E-02

LABCORE Completed Worklist Report for Worklist# 27144

Analyst: glt

Instrument: GEA03

Book#: 28B57

Method: LA-548-121

Rev/Mod

F-0

Worklist Comment: BY105, @GEA-01, tdm

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@GEA-01 C060-02 SOLID		4.501e-03	4.68e-03	103.977 % Recovery	
1 STD	0	@GEA-01 C060-02E SOLID		1	2.86	2.860 % Ct Error	
1 STD	0	@GEA-01 CS13702 SOLID		7.445e-03	7.47e-03	100.336 % Recovery	
1 STD	0	@GEA-01 CS13702E SOLID		1	2.75	2.750 % Ct Error	
2 BLNK-PREP	0	@GEA-01 C060-02 SOLID		1	<9.21e-3	uCi/g	
2 BLNK-PREP	0	@GEA-01 CS13702 SOLID		1	<1.39e-2	uCi/g	
3 SAMPLE	S98T002988	0 F @GEA-01 C060-02 SOLID		N/A	< 8.550e-03	8.55e-003 uCi/g	
3 SAMPLE	S98T002988	0 F @GEA-01 C060-02E SOLID		N/A	n/a	% Ct. Error	
3 SAMPLE	S98T002988	0 F @GEA-01 CS13702 SOLID		N/A	4.378e+01	uCi/g	
3 SAMPLE	S98T002988	0 F @GEA-01 CS13702E SOLID		N/A	0.430	% Ct. Error	
4 DUP	S98T002988	0 F @GEA-01 C060-02 SOLID		<8.55e-3	<8.49e-3	RPD	
4 DUP	S98T002988	0 F @GEA-01 C060-02E SOLID		1	n/a	% Ct Error	
4 DUP	S98T002988	0 F @GEA-01 CS13702 SOLID		4.38e+01	3.71e+01	16.564 RPD	
4 DUP	S98T002988	0 F @GEA-01 CS13702E SOLID		1.00	0.460	0.460 % Ct Error	
5 SAMPLE	S98T002989	0 F @GEA-01 C060-02 SOLID		N/A	< 7.627e-03	7.63e-003 uCi/g	
5 SAMPLE	S98T002989	0 F @GEA-01 C060-02E SOLID		N/A	n/a	% Ct. Error	
5 SAMPLE	S98T002989	0 F @GEA-01 CS13702 SOLID		N/A	5.547e+01	uCi/g	
5 SAMPLE	S98T002989	0 F @GEA-01 CS13702E SOLID		N/A	0.380	% Ct. Error	
6 DUP	S98T002989	0 F @GEA-01 C060-02 SOLID		<7.63e-3	<8.28e-3	RPD	
6 DUP	S98T002989	0 F @GEA-01 C060-02E SOLID		1	n/a	% Ct Error	
6 DUP	S98T002989	0 F @GEA-01 CS13702 SOLID		5.55e+01	5.64e+01	1.609 RPD	
6 DUP	S98T002989	0 F @GEA-01 CS13702E SOLID		1.00	0.370	0.370 % Ct Error	

Comments Section:

Comments for sample# S98T002988 and test @GEA-01 .
DL=0 => n/a.

Comments for sample# S98T002989 and test @GEA-01 .
DL=0 => n/a.

Final page for worklist# 27144

Analyst Signature

Date

Analyst Signature

Date

12-2-98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27144

Analyst: gll Instrument: GEA00 7 Book# 28657

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105, @GEA-01, tdm

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T002988	0	F	@GEA-01	SOLID	98000509	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T002988	0	F	@GEA-01	SOLID		
5	SAMPLE	S98T002989	0	F	@GEA-01	SOLID	98000509	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6	DUP	S98T002989	0	F	@GEA-01	SOLID		

Final page for worklist # 27144

gll 11-30-98
Analyst Signature Date

MLB 12/2/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 15:21:43.12 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27144
 Sample ID: WL27144-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3954.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:02.76 sec
 Dead Time: 0.1%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 14:31:09.31
 Decayed to: 30-NOV-1998 14:31:09.31
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.71*	6611	1.61	1323.59	1315	18	2.8		CS-137	7.47
0	1173.09	2965	1.80	2346.33	2336	20	4.2		CO-60	4.68
0	1332.32	2679	2.09	2664.86	2656	16	3.9		CO-60	4.69

Summary of Nuclide Activity

Page : 2

Sample ID : WL27144-STD

Acquisition date : 30-NOV-1998 14:31:09

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected	Decay Corr			
			uCi/L	uCi/L	2-Sigma Error	%Error	
CO-60	5.27Y	1.00	4.683E+00	4.683E+00	0.134E+00	2.86	
CS-137	30.00Y	1.00	7.472E+00	7.472E+00	0.206E+00	2.75	
Total Activity :			1.215E+01	1.215E+01			

Grand Total Activity :	1.215E+01	1.215E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27144-STD

Acquisition date : 30-NOV-1998 14:31:09

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	357.	477.59	6.4118E-01
NA-24	19.	1368.55	4.1604E-02
K-40	152.	1460.75	1.0568E+00
AR-41	24.	1293.60	5.0718E-02
SC-46	153.	1120.55	9.1174E-02
CR-51	354.	320.08	4.8070E-01
MN-54	145.	834.83	6.9073E-02
CO-56	165.	846.76	7.4295E-02
CO-58	153.	810.78	6.9446E-02
FE-59	146.	1099.25	1.5511E-01
ZN-65	147.	1115.55	1.7578E-01
SE-75	380.	264.66	7.4602E-02
KR-85	262.	514.00	1.3975E+01
SR-85	262.	514.01	6.3172E-02
Y-88	4.	1836.06	2.7082E-02
Y-91	41.	1204.67	1.7460E+01
NB-94	183.	871.09	8.0104E-02
NB-95	131.	765.78	6.1280E-02
ZR-95	141.	756.73	1.1496E-01
ZRNB-95	143.	724.18	2.7882E-01
RU-103	243.	497.08	6.2598E-02
RURH-106	157.	621.93	1.1223E+00
AG-108m	132.	722.94	6.4330E-02
CD-109	382.	88.03	1.1539E+00
AG-110M	294.	657.76	8.3508E-02
SN-113	323.	391.69	8.4365E-02
TE-123m	403.	159.00	3.6797E-02
SB-124	176.	602.73	5.8639E-02
SB-125	323.	427.89	1.9699E-01
TE-125m	413.	109.27	1.1775E+01
I-129	411.	39.60	0.0000E+00
I-131	309.	364.48	6.1766E-02
XE-131m	414.	163.93	1.6212E+00
BA-133	352.	356.02	8.4415E-02
CS-134	163.	604.70	5.6735E-02
CS-136	133.	818.51	6.5258E-02
CS-138	17.	1435.86	8.6894E-02
CE-139	405.	165.85	3.9517E-02
BA-140	189.	537.31	2.2072E-01
LA-140	9.	1596.21	3.4874E-02
CE-141	394.	145.44	6.1753E-02
CE-144	404.	133.51	2.7090E-01
CEPR-144	404.	133.51	5.4142E-01
EU-152	361.	121.78	1.0265E-01
EU-154	31.	1274.51	1.3927E-01
EU-155	398.	86.54	1.3339E-01
HF-181	279.	482.18	7.3830E-02
TA-182	349.	67.75	1.8795E-01
HG-203	393.	279.20	5.6536E-02
BI-207	194.	569.70	5.8558E-02
TL-208	383.	277.36	7.1732E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27144-STD

Acquisition date : 30-NOV-1998 14:31:09

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	405.	46.50	0.0000E+00
BI-212	159.	727.18	9.6432E-01
PB-212	486.	238.63	1.0451E-01
BI-214	207.	609.31	1.3942E-01
PB-214	403.	351.92	2.7636E-01
RA-224	473.	240.99	1.1453E+00
RA-226	467.	186.10	1.0395E+00
AC-228	184.	911.21	3.1352E-01
TH-228	380.	84.37	3.5445E+00
TH-229	388.	88.47	1.6833E-01
U-232	413.	57.78	9.7859E+01
PA-233	356.	312.17	1.2348E-01
UTH-233	423.	245.34	3.9558E+01
PA-234M	171.	1001.03	1.3437E+01
TH-234	314.	63.29	2.5775E+00
U-235	469.	185.71	6.3291E-02
NP-237	399.	86.48	3.5393E-01
U-237	387.	59.54	4.4287E-01
NP-238	201.	984.45	3.3644E-01
NP-239	353.	106.12	1.3847E-01
PU-239	407.	129.30	4.7261E+02
AM-241	387.	59.54	4.2477E-01
AM-243	336.	74.67	9.3285E-02

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 17:56:22.68 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27144
 Sample ID: WL27144-BLNK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Newing LT

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3955.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:02.23 sec
 Dead Time: 0.0%

Verified by:

Jones 12-2-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 15:25:51.82
 Decayed to: 30-NOV-1998 15:25:51.82
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	351.97*	90	1.64	704.33	699	10 56.4			
0	661.52*	74	1.64	1323.22	1316	15 78.6		CS-137	2.770E-02
0	1764.20*	19	2.00	3529.02	3522	16 94.1			

Summary of Nuclide Activity
Sample ID : WL27144-BLNK

Page : 2
Acquisition date : 30-NOV-1998 15:25:51

Total number of lines in spectrum 3
Number of unidentified lines 0
Number of lines tentatively identified by NID 3 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error Flags
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~2.770E-02~~	~~2.770E-02~~	~~2.177E-02~~	~~78.59~~
Total Activity :			2.770E-02	2.770E-02		12-2-98

Grand Total Activity : 2.770E-02 2.770E-02

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27144-BLNK

Acquisition date : 30-NOV-1998 15:25:51

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	92.	477.59	1.1190E-01
NA-24	20.	1368.55	1.4739E-02
K-40	364.	1460.75	5.3645E-01
AR-41	21.	1293.60	2.1360E-02
SC-46	75.	1120.55	2.1654E-02
CR-51	192.	320.08	1.1933E-01
MN-54	57.	834.83	1.4774E-02
CO-56	61.	846.76	1.5541E-02
CO-58	44.	810.78	1.2933E-02
FE-59	43.	1099.25	2.9273E-02
CO-60	38.	1332.50	1.8292E-02
ZN-65	47.	1115.55	3.4340E-02
SE-75	221.	264.66	1.9123E-02
KR-85	176.	514.00	3.8501E+00
SR-85	176.	514.01	1.7392E-02
Y-88	6.	1836.06	1.0464E-02
Y-91	42.	1204.67	5.8926E+00
NB-94	46.	871.09	1.3968E-02
NB-95	71.	765.78	1.5271E-02
ZR-95	62.	756.73	2.6043E-02
ZRNB-95	63.	724.18	6.3201E-02
RU-103	89.	497.08	1.2907E-02
RURH-106	83.	621.93	2.7651E-01
AG-108m	58.	722.94	1.4515E-02
CD-109	236.	88.03	3.0450E-01
AG-110M	69.	657.76	1.3984E-02
SN-113	122.	391.69	1.7606E-02
TE-123m	260.	159.00	9.9131E-03
SB-124	81.	602.73	1.3502E-02
SB-125	141.	427.89	4.4083E-02
TE-125m	230.	109.27	2.9553E+00
I-129	392.	39.60	0.0000E+00
I-131	141.	364.48	1.4166E-02
XE-131m	272.	163.93	4.4169E-01
BA-133	137.	356.02	1.7870E-02
CS-134	70.	604.70	1.2695E-02
CS-136	53.	818.51	1.4210E-02
CS-138	21.	1435.86	6.4892E-02
CE-139	249.	165.85	1.0423E-02
BA-140	76.	537.31	4.7711E-02
LA-140	12.	1596.21	1.3701E-02
CE-141	262.	145.44	1.6896E-02
CE-144	230.	133.51	6.8824E-02
CEPR-144	230.	133.51	1.3755E-01
EU-152	263.	121.78	2.9335E-02
EU-154	31.	1274.51	4.6722E-02
EU-155	278.	86.54	3.7367E-02
HF-181	105.	482.18	1.5438E-02
TA-182	223.	67.75	5.0439E-02
HG-203	175.	279.20	1.2749E-02
BI-207	79.	569.70	1.2714E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27144-BLKN

Acquisition date : 30-NOV-1998 15:25:51

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	197.	277.36	1.7327E-01
PB-210	441.	46.50	0.0000E+00
BI-212	85.	727.18	2.3876E-01
PB-212	266.	238.63	2.5996E-02
BI-214	148.	609.31	3.9665E-02
PB-214	256.	351.92	1.6425E-01
RA-224	265.	240.99	2.9004E-01
RA-226	258.	186.10	2.5988E-01
AC-228	94.	911.21	7.5843E-02
TH-228	265.	84.37	9.9139E-01
TH-229	230.	88.47	4.3642E-02
U-232	405.	57.78	3.2304E+01
PA-233	170.	312.17	2.8871E-02
UTH-233	200.	245.34	9.1751E+00
PA-234M	33.	1001.03	2.0851E+00
TH-234	204.	63.29	6.9731E-01
U-235	254.	185.71	1.5669E-02
NP-237	277.	86.48	9.8790E-02
U-237	308.	59.54	1.3267E-01
NP-238	42.	984.45	5.4470E-02
NP-239	244.	106.12	3.9012E-02
PU-239	276.	129.30	1.3046E+02
AM-241	308.	59.54	1.2679E-01
AM-243	268.	74.67	2.7841E-02

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 20:50:44.87 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27144
 Sample ID: S98T002988-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3956.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:30.80 sec
 Dead Time: 0.3%

Verified by:

12-2-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 18:19:44.93
 Decayed to: 30-NOV-1998 18:19:44.93
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.68*	230982	1.59	1323.53	1315	18	0.4		CS-137	87.0
0	1120.31*	31	1.93	2240.75	2236	14	89.7			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002988-SAM

Acquisition date : 30-NOV-1998 18:19:44

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
CS-137	30.00Y	1.00	8.702E+01	8.702E+01			0.037E+01	0.43	
Total Activity :			8.702E+01	8.702E+01					

Grand Total Activity : 8.702E+01 8.702E+01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002988-SAM

Acquisition date : 30-NOV-1998 18:19:44

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	7364.	477.59	9.4917E-01
NA-24	20.	1368.55	1.4937E-02
K-40	360.	1460.75	5.3312E-01
AR-41	29.	1293.60	2.4551E-02
SC-46	76.	1120.55	2.1816E-02
CR-51	5989.	320.08	6.4492E-01
MN-54	64.	834.83	1.5610E-02
CO-56	55.	846.76	1.4816E-02
CO-58	71.	810.78	1.6061E-02
FE-59	44.	1099.25	2.9497E-02
CO-60	32.	1332.50	1.6991E-02
ZN-65	39.	1115.55	3.1447E-02
SE-75	6781.	264.66	1.0269E-01
KR-85	3874.	514.00	1.7443E+01
SR-85	3872.	514.01	7.8884E-02
Y-88	8.	1836.06	1.1902E-02
Y-91	44.	1204.67	5.9847E+00
NB-94	58.	871.09	1.5552E-02
NB-95	61.	765.78	1.4246E-02
ZR-95	78.	756.73	2.8918E-02
ZRNB-95	87.	724.18	7.3436E-02
RU-103	4620.	497.08	8.8444E-02
RURH-106	1419.	621.93	1.0927E+00
AG-108m	91.	722.94	1.7970E-02
CD-109	5732.	88.03	1.4573E+00
AG-110M	6007.	657.76	1.2262E-01
SN-113	6453.	391.69	1.2265E-01
TE-123m	6790.	159.00	4.9257E-02
SB-124	1441.	602.73	5.4402E-02
SB-125	7500.	427.89	3.0845E-01
TE-125m	5966.	109.27	1.4617E+01
I-129	5575.	39.60	0.0000E+00
I-131	6067.	364.48	8.9181E-02
XE-131m	6840.	163.93	2.1543E+00
BA-133	6068.	356.02	1.1418E-01
CS-134	1431.	604.70	5.4448E-02
CS-136	56.	818.51	1.4555E-02
CS-138	19.	1435.86	6.2101E-02
CE-139	6863.	165.85	5.3105E-02
BA-140	2657.	537.31	2.6797E-01
LA-140	18.	1596.21	1.5937E-02
CE-141	6653.	145.44	8.2836E-02
CE-144	6763.	133.51	3.6166E-01
CEPR-144	6764.	133.51	7.2269E-01
EU-152	6423.	121.78	1.4098E-01
EU-154	38.	1274.51	5.1344E-02
EU-155	5734.	86.54	1.6528E-01
HF-181	6006.	482.18	1.1115E-01
TA-182	5430.	67.75	2.4171E-01
HG-203	6428.	279.20	7.4596E-02
BI-207	1588.	569.70	5.4338E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002988-SAM

Acquisition date : 30-NOV-1998 18:19:44

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	6390.	277.36	9.5501E-01
PB-210	5682.	46.50	0.0000E+00
BI-212	93.	727.18	2.4933E-01
PB-212	7675.	238.63	1.3579E-01
BI-214	1526.	609.31	1.2319E-01
PB-214	6170.	351.92	7.8609E-01
RA-224	7601.	240.99	1.5104E+00
RA-226	8778.	186.10	1.4714E+00
AC-228	95.	911.21	7.6303E-02
TH-228	5899.	84.37	4.5527E+00
TH-229	5722.	88.47	2.1104E-01
U-232	5534.	57.78	1.1698E+02
PA-233	6016.	312.17	1.6558E-01
UTH-233	7135.	245.34	5.3014E+01
PA-234M	44.	1001.03	2.3682E+00
TH-234	5314.	63.29	3.4480E+00
U-235	8676.	185.71	8.8907E-02
NP-237	5746.	86.48	4.3822E-01
U-237	5514.	59.54	5.4751E-01
NP-238	48.	984.45	5.7829E-02
NP-239	5942.	106.12	1.8699E-01
PU-239	6633.	129.30	6.2284E+02
AM-241	5514.	59.54	5.2329E-01
AM-243	5675.	74.67	1.2477E-01

*****

* 222-S Laboratory Counting Room 30-NOV-1998 23:27:02.94 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27144
 Sample ID: S98T002988-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Newright

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3957.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:26.56 sec
 Dead Time: 0.3%

Verified by:

R. Jones 12-2-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 20:56:09.29
 Decayed to: 30-NOV-1998 20:56:09.29
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.69*	196843	1.58	1323.55	1315	18	0.5		CS-137	74.2
0	1119.99*	31	1.61	2240.11	2234	13	88.3			
0	1764.19*	13	2.35	3529.00	3522	121	24.2			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002988-DUP

Acquisition date : 30-NOV-1998 20:56:09

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected uCi/L	Decay Corr uCi/L			
CS-137	30.00Y	1.00	7.416E+01	7.416E+01	0.034E+01	0.46	
Total Activity :			7.416E+01	7.416E+01			

Grand Total Activity :	7.416E+01	7.416E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002988-DUP

Acquisition date : 30-NOV-1998 20:56:09

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	6420.	477.59	8.8670E-01
NA-24	22.	1368.55	1.5498E-02
K-40	388.	1460.75	5.5349E-01
AR-41	24.	1293.60	2.2606E-02
SC-46	77.	1120.55	2.1949E-02
CR-51	5108.	320.08	5.9594E-01
MN-54	59.	834.83	1.5077E-02
CO-56	70.	846.76	1.6488E-02
CO-58	54.	810.78	1.4170E-02
FE-59	41.	1099.25	2.8612E-02
CO-60	32.	1332.50	1.6964E-02
ZN-65	44.	1115.55	3.3432E-02
SE-75	5744.	264.66	9.4571E-02
KR-85	3232.	514.00	1.5946E+01
SR-85	3231.	514.01	7.2121E-02
Y-88	12.	1836.06	1.4227E-02
Y-91	49.	1204.67	6.3201E+00
NB-94	54.	871.09	1.4967E-02
NB-95	66.	765.78	1.4767E-02
ZR-95	77.	756.73	2.8795E-02
ZRNB-95	72.	724.18	6.7262E-02
RU-103	3924.	497.08	8.1575E-02
RURH-106	1143.	621.93	9.8225E-01
AG-108m	64.	722.94	1.5228E-02
CD-109	4924.	88.03	1.3516E+00
AG-110M	5013.	657.76	1.1210E-01
SN-113	5500.	391.69	1.1329E-01
TE-123m	5754.	159.00	4.5373E-02
SB-124	1319.	602.73	5.2084E-02
SB-125	6535.	427.89	2.8806E-01
TE-125m	5135.	109.27	1.3568E+01
I-129	4824.	39.60	0.0000E+00
I-131	5244.	364.48	8.2961E-02
XE-131m	5646.	163.93	1.9586E+00
BA-133	5155.	356.02	1.0530E-01
CS-134	1276.	604.70	5.1464E-02
CS-136	60.	818.51	1.4947E-02
CS-138	13.	1435.86	5.2259E-02
CE-139	5737.	165.85	4.8584E-02
BA-140	2223.	537.31	2.4535E-01
LA-140	21.	1596.21	1.7385E-02
CE-141	5712.	145.44	7.6799E-02
CE-144	5746.	133.51	3.3355E-01
CEPR-144	5745.	133.51	6.6647E-01
EU-152	5585.	121.78	1.3153E-01
EU-154	24.	1274.51	4.1714E-02
EU-155	5013.	86.54	1.5462E-01
HF-181	5162.	482.18	1.0311E-01
TA-182	4684.	67.75	2.2462E-01
HG-203	5363.	279.20	6.8185E-02
BI-207	1444.	569.70	5.1854E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002988-DUP

Acquisition date : 30-NOV-1998 20:56:09

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	5433.	277.36	8.8108E-01
PB-210	4845.	46.50	0.0000E+00
BI-212	87.	727.18	2.4137E-01
PB-212	6465.	238.63	1.2471E-01
BI-214	1305.	609.31	1.1407E-01
PB-214	5154.	351.92	7.1862E-01
RA-224	6335.	240.99	1.3798E+00
RA-226	7494.	186.10	1.3602E+00
AC-228	86.	911.21	7.2760E-02
TH-228	4935.	84.37	4.1669E+00
TH-229	4920.	88.47	1.9580E-01
U-232	4723.	57.78	1.0814E+02
PA-233	5304.	312.17	1.5555E-01
UTH-233	6202.	245.34	4.9451E+01
PA-234M	47.	1001.03	2.4442E+00
TH-234	4669.	63.29	3.2337E+00
U-235	7414.	185.71	8.2231E-02
NP-237	5012.	86.48	4.0951E-01
U-237	4640.	59.54	5.0261E-01
NP-238	49.	984.45	5.8282E-02
NP-239	5144.	106.12	1.7408E-01
PU-239	5667.	129.30	5.7602E+02
AM-241	4640.	59.54	4.8037E-01
AM-243	4726.	74.67	1.1395E-01

*****

* 222-S Laboratory Counting Room 1-DEC-1998 06:36:26.87 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27144
 Sample ID: S98T002989-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3960.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:39.00 sec
 Dead Time: 0.4%

Verified by:

12-2-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 1-DEC-1998 04:05:19.13
 Decayed to: 1-DEC-1998 04:05:19.13
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.68*	296907	1.59	1323.54	1315	18	0.4		CS-137	112.
0	910.82*	37	3.47	1821.75	1816	13	90.0			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002989-SAM

Acquisition date : 1-DEC-1998 04:05:19

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.119E+02	1.119E+02	0.004E+02	0.38	
Total Activity :			1.119E+02	1.119E+02			

Grand Total Activity :	1.119E+02	1.119E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002989-SAM

Acquisition date : 1-DEC-1998 04:05:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	9647.	477.59	1.0855E+00
NA-24	21.	1368.55	1.5252E-02
K-40	369.	1460.75	5.3955E-01
AR-41	36.	1293.60	2.7221E-02
SC-46	55.	1120.55	1.8824E-02
CR-51	7605.	320.08	7.2613E-01
MN-54	64.	834.83	1.5655E-02
CO-56	74.	846.76	1.6939E-02
CO-58	75.	810.78	1.6570E-02
FE-59	48.	1099.25	3.0730E-02
CO-60	26.	1332.50	1.5400E-02
ZN-65	47.	1115.55	3.4159E-02
SE-75	8364.	264.66	1.1397E-01
KR-85	4765.	514.00	1.9329E+01
SR-85	4764.	514.01	8.7424E-02
Y-88	8.	1836.06	1.2020E-02
Y-91	43.	1204.67	5.9679E+00
NB-94	67.	871.09	1.6575E-02
NB-95	87.	765.78	1.6805E-02
ZR-95	73.	756.73	2.8103E-02
ZRNB-95	102.	724.18	7.9172E-02
RU-103	5844.	497.08	9.9382E-02
RURH-106	1880.	621.93	1.2550E+00
AG-108m	97.	722.94	1.8524E-02
CD-109	7377.	88.03	1.6518E+00
AG-110M	7709.	657.76	1.3879E-01
SN-113	8453.	391.69	1.4023E-01
TE-123m	8447.	159.00	5.4901E-02
SB-124	1894.	602.73	6.2250E-02
SB-125	9687.	427.89	3.5027E-01
TE-125m	7845.	109.27	1.6745E+01
I-129	7140.	39.60	0.0000E+00
I-131	7864.	364.48	1.0144E-01
XE-131m	8453.	163.93	2.3931E+00
BA-133	7789.	356.02	1.2924E-01
CS-134	1824.	604.70	6.1355E-02
CS-136	76.	818.51	1.6796E-02
CS-138	22.	1435.86	6.6373E-02
CE-139	8554.	165.85	5.9243E-02
BA-140	3327.	537.31	2.9949E-01
LA-140	12.	1596.21	1.3702E-02
CE-141	8651.	145.44	9.4376E-02
CE-144	8621.	133.51	4.0800E-01
CEPR-144	8621.	133.51	8.1527E-01
EU-152	8268.	121.78	1.5982E-01
EU-154	40.	1274.51	5.2390E-02
EU-155	7470.	86.54	1.8846E-01
HF-181	7703.	482.18	1.2577E-01
TA-182	6778.	67.75	2.6981E-01
HG-203	8213.	279.20	8.4248E-02
BI-207	2088.	569.70	6.2195E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002989-SAM

Acquisition date : 1-DEC-1998 04:05:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	8267.	277.36	1.0853E+00
PB-210	6960.	46.50	0.0000E+00
BI-212	101.	727.18	2.5921E-01
PB-212	9671.	238.63	1.5232E-01
BI-214	1968.	609.31	1.3969E-01
PB-214	7708.	351.92	8.7865E-01
RA-224	9529.	240.99	1.6900E+00
RA-226	11299.	186.10	1.6681E+00
AC-228	107.	911.21	8.0504E-02
TH-228	7277.	84.37	5.0529E+00
TH-229	7364.	88.47	2.3919E-01
U-232	6968.	57.78	1.3115E+02
PA-233	7706.	312.17	1.8723E-01
UTH-233	9280.	245.34	6.0411E+01
PA-234M	58.	1001.03	2.6896E+00
TH-234	6819.	63.29	3.9022E+00
U-235	11206.	185.71	1.0097E-01
NP-237	7458.	86.48	4.9879E-01
U-237	6886.	59.54	6.1137E-01
NP-238	50.	984.45	5.8777E-02
NP-239	7668.	106.12	2.1224E-01
PU-239	8577.	129.30	7.0762E+02
AM-241	6886.	59.54	5.8432E-01
AM-243	7120.	74.67	1.3963E-01

*****
 * 222-S Laboratory Counting Room 1-DEC-1998 09:14:16.82 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27144
 Sample ID: S98T2989-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3961.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:39.38 sec
 Dead Time: 0.4%

Verified by:

12-2-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 1-DEC-1998 06:43:08.57
 Decayed to: 1-DEC-1998 06:43:08.57
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.66*	301443	1.60	1323.51	1315	18	0.4		CS-137	114.

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2989-DUP

Acquisition date : 1-DEC-1998 06:43:08

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma
			Uncorrected	Decay Corr		
			uCi/L	uCi/L	2-Sigma Error	%Error
CS-137	30.00Y	1.00	1.136E+02	1.136E+02	0.004E+02	0.37
Total Activity :			1.136E+02	1.136E+02		

Grand Total Activity :	1.136E+02	1.136E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Sample ID : S98T2989-DUP

Acquisition date : 1-DEC-1998 06:43:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	9633.	477.59	1.0847E+00
NA-24	26.	1368.55	1.6679E-02
K-40	399.	1460.75	5.6104E-01
AR-41	33.	1293.60	2.6014E-02
SC-46	73.	1120.55	2.1455E-02
CR-51	7777.	320.08	7.3421E-01
MN-54	66.	834.83	1.5859E-02
CO-56	65.	846.76	1.5971E-02
CO-58	87.	810.78	1.7674E-02
FE-59	39.	1099.25	2.7847E-02
CO-60	31.	1332.50	1.6750E-02
ZN-65	47.	1115.55	3.4378E-02
SE-75	8536.	264.66	1.1513E-01
KR-85	4796.	514.00	1.9391E+01
SR-85	4794.	514.01	8.7691E-02
Y-88	10.	1836.06	1.2989E-02
Y-91	46.	1204.67	6.1350E+00
NB-94	63.	871.09	1.6067E-02
NB-95	66.	765.78	1.4751E-02
ZR-95	69.	756.73	2.7296E-02
ZRNB-95	78.	724.18	6.9815E-02
RU-103	5971.	497.08	1.0045E-01
RURH-106	1877.	621.93	1.2543E+00
AG-108m	81.	722.94	1.7016E-02
CD-109	7469.	88.03	1.6620E+00
AG-110M	7988.	657.76	1.4126E-01
SN-113	8320.	391.69	1.3913E-01
TE-123m	8651.	159.00	5.5554E-02
SB-124	1866.	602.73	6.1790E-02
SB-125	9788.	427.89	3.5207E-01
TE-125m	7793.	109.27	1.6690E+01
I-129	7371.	39.60	0.0000E+00
I-131	7921.	364.48	1.0181E-01
XE-131m	8650.	163.93	2.4206E+00
BA-133	7832.	356.02	1.2960E-01
CS-134	1847.	604.70	6.1736E-02
CS-136	72.	818.51	1.6363E-02
CS-138	24.	1435.86	6.9081E-02
CE-139	8522.	165.85	5.9132E-02
BA-140	3388.	537.31	3.0221E-01
LA-140	14.	1596.21	1.4244E-02
CE-141	8425.	145.44	9.3145E-02
CE-144	8574.	133.51	4.0688E-01
CEPR-144	8573.	133.51	8.1299E-01
EU-152	8499.	121.78	1.6202E-01
EU-154	34.	1274.51	4.8445E-02
EU-155	7399.	86.54	1.8757E-01
HF-181	7822.	482.18	1.2673E-01
TA-182	6810.	67.75	2.7044E-01
HG-203	8192.	279.20	8.4141E-02
BI-207	2105.	569.70	6.2447E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2989-DUP

Acquisition date : 1-DEC-1998 06:43:08

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	8308.	277.36	1.0879E+00
PB-210	7100.	46.50	0.0000E+00
BI-212	88.	727.18	2.4293E-01
PB-212	9757.	238.63	1.5300E-01
BI-214	1919.	609.31	1.3795E-01
PB-214	7751.	351.92	8.8113E-01
RA-224	9594.	240.99	1.6957E+00
RA-226	11459.	186.10	1.6798E+00
AC-228	109.	911.21	8.1394E-02
TH-228	7297.	84.37	5.0598E+00
TH-229	7459.	88.47	2.4072E-01
U-232	7007.	57.78	1.3152E+02
PA-233	7830.	312.17	1.8872E-01
UTH-233	9359.	245.34	6.0668E+01
PA-234M	51.	1001.03	2.5288E+00
TH-234	6649.	63.29	3.8536E+00
U-235	11315.	185.71	1.0145E-01
NP-237	7399.	86.48	4.9683E-01
U-237	6833.	59.54	6.0901E-01
NP-238	51.	984.45	5.9560E-02
NP-239	7510.	106.12	2.1005E-01
PU-239	8595.	129.30	7.0836E+02
AM-241	6833.	59.54	5.8206E-01
AM-243	7278.	74.67	1.4116E-01

LABCORE Completed Worklist Report for Worklist# 27168

Analyst: gll

Instrument: GEA04

Book#: 28B57

Method: LA-548-121

Rev/Mod F-0

Worklist Comment: BY105 (2), @GEA-01, SS= 1.0mL, STD= 1.0mL skm

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@GEA-01 C060-02 SOLID		4.521e-03	4.54e-03	100.420 % Recovery	
1 STD	0	@GEA-01 C060-02E SOLID		1	1.61	1.610 % Ct Error	
1 STD	0	@GEA-01 CS13702 SOLID		7.450e-03	7.61e-03	102.148 % Recovery	
1 STD	0	@GEA-01 CS13702E SOLID		1	1.49	1.490 % Ct Error	
2 BLNK-PREP	0	@GEA-01 C060-02 SOLID		1	<4.68e-3		uCi/g
2 BLNK-PREP	0	@GEA-01 CS13702 SOLID		1	<7.68e-3		uCi/g
3 SAMPLE	S98T002598	0 F @GEA-01 C060-02 SOLID		N/A	4.566e-03	4.57e-003	uCi/g
3 SAMPLE	S98T002598	0 F @GEA-01 C060-02E SOLID		N/A	n/a		% Ct. Error
3 SAMPLE	S98T002598	0 F @GEA-01 CS13702 SOLID		N/A	3.372e+00		uCi/g
3 SAMPLE	S98T002598	0 F @GEA-01 CS13702E SOLID		N/A	0.870		% Ct. Error
4 DUP	S98T002598	0 F @GEA-01 C060-02 SOLID		<4.57e-3	<4.88e-3		RFD
4 DUP	S98T002598	0 F @GEA-01 C060-02E SOLID		1	n/a		% Ct Error
4 DUP	S98T002598	0 F @GEA-01 CS13702 SOLID		3.37e+00	4.30e+00	24.250 RFD	
4 DUP	S98T002598	0 F @GEA-01 CS13702E SOLID		1.00	0.760	0.760	% Ct Error

Comments Section:

Comments for sample# S98T002598 and test @GEA-01 .

DL=0 => n/a.

DL=0 => n/a.

DATA
NOT USED

Final page for worklist# 27168

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

[Signature] 11-19-98
Reviewer Signature Date

S98T002598 reanalyzed on WL# 27288. [Signature]
3/4/98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27168

Analyst: gll Instrument: GEA00 4 Book# 28857

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 (2), @GEA-01, SS= 1.0mL, STD= 1.0mL skm Run

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	SOLID		
2 BLNK-PREP			@GEA-01	SOLID		
3 SAMPLE	S98T002598	0 F	@GEA-01	SOLID	98000436	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4 DUP	S98T002598	0 F	@GEA-01	SOLID		

Final page for worklist # 27168

gll 11-18-98
Analyst Signature Date

MRB 11/19/98
Analyst Signature Date

DATA
NOT USED
11/23/99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

4539

*****
 * 222-S Laboratory Counting Room 18-NOV-1998 12:07:43.96 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27168
 Sample ID: WL27168-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Newlight

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4382.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:09.49 sec
 Dead Time: 0.3%

Verified by:

RJm 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 11:17:03.90
 Decayed to: 18-NOV-1998 11:17:03.90
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	59.45	207	1.10	119.04	115	8	63.5		AM-241	0.252
0	661.53*	22539	1.61	1323.12	1314	18	1.5		CS-137	7.61
0	1173.13	9858	1.98	2346.36	2339	18	2.3		CO-60	4.53
0	1332.37	8915	2.08	2664.88	2654	22	2.2		CO-60	4.54

DATA
 NOT USED

Summary of Nuclide Activity
Sample ID : WL27168-STD

HNF-1663 REV. 0

Page : 2
Acquisition date : 18-NOV-1998 11:17:03

Total number of lines in spectrum 4
Number of unidentified lines 0
Number of lines tentatively identified by NID 4 100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.540E+00	4.540E+00	0.073E+00	1.61	
CS-137	30.00Y	1.00	7.607E+00	7.607E+00	0.113E+00	1.49	
~~AM-241~~	~~432.20Y~~	~~1.00~~	~~2.521E-01~~	~~2.521E-01~~	~~1.601E-01~~	~~63.51~~	~~RAJ~~
Total Activity :			1.240E+01	1.240E+01			11-19-98

Grand Total Activity : 1.240E+01 1.240E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27168-STD

Acquisition date : 18-NOV-1998 11:17:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1209.	477.59	3.3554E-01
NA-24	62.	1368.55	2.0883E-02
K-40	766.	1460.75	6.7902E-01
AR-41	87.	1293.60	2.6899E-02
SC-46	498.	1120.55	4.7208E-02
CR-51	1168.	320.08	2.2546E-01
MN-54	579.	834.83	4.0175E-02
CO-56	596.	846.76	4.1264E-02
CO-58	563.	810.78	3.8907E-02
FE-59	573.	1099.25	8.8129E-02
ZN-65	524.	1115.55	9.5025E-02
SE-75	1360.	264.66	3.4832E-02
KR-85	908.	514.00	7.4765E+00
SR-85	908.	514.01	3.3791E-02
Y-88	23.	1836.06	1.7343E-02
Y-91	181.	1204.67	1.0212E+01
NB-94	594.	871.09	4.2156E-02
NB-95	497.	765.78	3.4805E-02
ZR-95	491.	756.73	6.2731E-02
ZRNB-95	498.	724.18	1.5210E-01
RU-103	916.	497.08	3.4660E-02
RURH-106	604.	621.93	6.4159E-01
AG-108m	492.	722.94	3.6210E-02
CD-109	1052.	88.03	5.6195E-01
AG-110M	1507.	657.76	5.5414E-02
SN-113	1195.	391.69	4.4121E-02
TE-123m	1292.	159.00	1.7278E-02
SB-124	612.	602.73	3.1811E-02
SB-125	1196.	427.89	1.0518E-01
TE-125m	1188.	109.27	5.6089E+00
I-129	1097.	39.60	1.0759E+01
I-131	1205.	364.48	3.2467E-02
XE-131m	1267.	163.93	7.3888E-01
BA-133	1094.	356.02	3.9503E-02
CS-134	616.	604.70	3.2077E-02
CS-136	590.	818.51	4.0036E-02
CS-138	47.	1435.86	4.0062E-02
CE-139	1239.	165.85	1.7969E-02
BA-140	690.	537.31	1.2150E-01
LA-140	23.	1596.21	1.5825E-02
CE-141	1301.	145.44	3.0011E-02
CE-144	1293.	133.51	1.3171E-01
CEPR-144	1293.	133.51	2.6321E-01
EU-152	1199.	121.78	5.1510E-02
EU-154	111.	1274.51	7.3827E-02
EU-155	1152.	86.54	6.6894E-02
HF-181	1102.	482.18	4.1628E-02
TA-182	1099.	67.75	1.0085E-01
HG-203	1279.	279.20	2.5521E-02
BI-207	614.	569.70	3.0235E-02
TL-208	1265.	277.36	3.2555E-01

DATA
NOT USED

4542

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27168-STD

Acquisition date : 18-NOV-1998 11:17:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1071.	46.50	6.2581E+00
BI-212	514.	727.18	5.0777E-01
PB-212	1562.	238.63	4.5720E-02
BI-214	787.	609.31	7.9347E-02
PB-214	1260.	351.92	1.2948E-01
RA-224	1489.	240.99	4.9600E-01
RA-226	1567.	186.10	4.8173E-01
AC-228	784.	911.21	1.8797E-01
TH-228	1223.	84.37	1.8819E+00
TH-229	1056.	88.47	8.1508E-02
U-232	1138.	57.78	3.9919E+01
PA-233	1197.	312.17	5.8116E-02
UTH-233	1400.	245.34	1.7571E+01
PA-234M	621.	1001.03	7.3937E+00
TH-234	1082.	63.29	1.3935E+00
U-235	1556.	185.71	2.9193E-02
NP-237	1159.	86.48	1.7770E-01
U-237	1238.	59.54	2.1146E-01
NP-238	579.	984.45	1.6589E-01
NP-239	1155.	106.12	7.0551E-02
PU-239	1251.	129.30	2.2654E+02
AM-243	1162.	74.67	5.2451E-02

DATA
NOT USED

*****

* 222-S Laboratory Counting Room 18-NOV-1998 15:17:39.59 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27168
 Sample ID: WL27168-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by: .

new light

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4383.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:08.94 sec
 Dead Time: 0.1%

Verified by:

RAGme 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 12:46:59.12
 Decayed to: 18-NOV-1998 12:46:59.12
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	29.24*	198	1.43	58.64	56	7 87.7		I-129	10.5
0	92.70*	192	1.23	185.54	180	12 88.8	1.13E+01		

DATA
NOT USED

Summary of Nuclide Activity

Page : 2

Sample ID : WL27168-BLK

Acquisition date : 18-NOV-1998 12:46:59

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
~~I-129~~	~~5.73E+09D~~	~~1.00~~	~~1.055E+01~~	~~1.055E+01~~	~~0.925E+01~~	~~87.69~~			~~RTJ~~
Total Activity :			1.055E+01	1.055E+01					11-19-98

Grand Total Activity : 1.055E+01 1.055E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27168-BLK

Acquisition date : 18-NOV-1998 12:46:59

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	491.	477.59	7.1966E-02
NA-24	116.	1368.55	9.7248E-03
K-40	2203.	1460.75	3.8067E-01
AR-41	143.	1293.60	1.5055E-02
SC-46	383.	1120.55	1.3856E-02
CR-51	894.	320.08	6.5968E-02
MN-54	300.	834.83	9.7250E-03
CO-56	254.	846.76	9.0916E-03
CO-58	239.	810.78	8.5606E-03
FE-59	205.	1099.25	1.7872E-02
CO-60	132.	1332.50	9.5390E-03
ZN-65	215.	1115.55	2.0583E-02
SE-75	1029.	264.66	1.0124E-02
KR-85	688.	514.00	2.1756E+00
SR-85	686.	514.01	9.8271E-03
Y-88	53.	1836.06	8.4348E-03
Y-91	229.	1204.67	3.8147E+00
NB-94	244.	871.09	9.1259E-03
NB-95	291.	765.78	8.9455E-03
ZR-95	331.	756.73	1.7269E-02
ZRNB-95	306.	724.18	4.0041E-02
RU-103	473.	497.08	8.3706E-03
RURH-106	342.	621.93	1.6214E-01
AG-108m	298.	722.94	9.4688E-03
CD-109	1024.	88.03	1.8491E-01
AG-110M	361.	657.76	9.1752E-03
SN-113	635.	391.69	1.0787E-02
TE-123m	1146.	159.00	5.4303E-03
SB-124	380.	602.73	8.4121E-03
SB-125	569.	427.89	2.4365E-02
TE-125m	979.	109.27	1.6999E+00
I-131	643.	364.48	7.9782E-03
XE-131m	1117.	163.93	2.3203E-01
BA-133	716.	356.02	1.0693E-02
CS-134	376.	604.70	8.4066E-03
CS-136	254.	818.51	8.8797E-03
CS-137	859.	661.66	1.5634E-02
CS-138	101.	1435.86	3.9396E-02
CE-139	1129.	165.85	5.7233E-03
BA-140	389.	537.31	3.0682E-02
LA-140	61.	1596.21	8.3555E-03
CE-141	1117.	145.44	9.2890E-03
CE-144	1024.	133.51	3.9160E-02
CEPR-144	1025.	133.51	7.8253E-02
EU-152	1004.	121.78	1.5736E-02
EU-154	166.	1274.51	2.9812E-02
EU-155	1096.	86.54	2.1756E-02
HF-181	519.	482.18	9.6031E-03
TA-182	827.	67.75	2.9254E-02
HG-203	948.	279.20	7.3447E-03
BI-207	418.	569.70	8.3586E-03

DATA
NOT USED

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27168-BLK

Acquisition date : 18-NOV-1998 12:46:59

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	992.	277.36	9.6263E-02
PB-210	805.	46.50	1.8188E+00
BI-212	358.	727.18	1.4203E-01
PB-212	1466.	238.63	1.4771E-02
BI-214	823.	609.31	2.7030E-02
PB-214	1066.	351.92	8.8284E-02
RA-224	1266.	240.99	1.5364E-01
RA-226	1254.	186.10	1.4391E-01
AC-228	500.	911.21	5.0311E-02
TH-228	1124.	84.37	6.0191E-01
TH-229	1004.	88.47	2.6510E-02
U-232	772.	57.78	1.0997E+01
PA-233	796.	312.17	1.5867E-02
UTH-233	1009.	245.34	4.9853E+00
PA-234M	222.	1001.03	1.4979E+00
TH-234	859.	63.29	4.1476E-01
U-235	1257.	185.71	8.7607E-03
NP-237	1106.	86.48	5.7898E-02
U-237	797.	59.54	5.6976E-02
NP-238	218.	984.45	3.4858E-02
NP-239	967.	106.12	2.1770E-02
PU-239	1043.	129.30	6.9047E+01
AM-241	797.	59.54	5.4459E-02
AM-243	1106.	74.67	1.7065E-02

DATA
NOT USED

*****

* 222-S Laboratory Counting Room 18-NOV-1998 18:27:17.08 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27168
 Sample ID: S98T002598-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4385.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:18.36 sec
 Dead Time: 0.2%

Verified by:

W Jones 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 15:56:27.63
 Decayed to: 18-NOV-1998 15:56:27.63
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.53*	61085	1.64	1323.12	1314	18	0.9		CS-137	6.87

DATA
NOT USED

Summary of Nuclide Activity

Sample ID : S98T002598-SAM

Page : 2

Acquisition date : 18-NOV-1998 15:56:27

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	6.872E+00	6.872E+00	0.060E+00	0.87	
Total Activity :			6.872E+00	6.872E+00			

Grand Total Activity :	6.872E+00	6.872E+00

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report
 Sample ID : S98T002598-SAM

Page : 3
 Acquisition date : 18-NOV-1998 15:56:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2468.	477.59	1.5911E-01
NA-24	80.	1368.55	8.1564E-03
K-40	2217.	1460.75	3.8180E-01
AR-41	147.	1293.60	1.5276E-02
SC-46	341.	1120.55	1.3091E-02
CR-51	2188.	320.08	1.0248E-01
MN-54	277.	834.83	9.3653E-03
CO-56	281.	846.76	9.5392E-03
CO-58	241.	810.78	8.5953E-03
FE-59	195.	1099.25	1.7436E-02
CO-60	125.	1332.50	9.3019E-03
ZN-65	218.	1115.55	2.0720E-02
SE-75	2524.	264.66	1.5759E-02
KR-85	1647.	514.00	3.3388E+00
SR-85	1645.	514.01	1.5097E-02
Y-88	59.	1836.06	8.8881E-03
Y-91	209.	1204.67	3.6511E+00
NB-94	265.	871.09	9.4936E-03
NB-95	326.	765.78	9.4574E-03
ZR-95	298.	756.73	1.6399E-02
ZRNB-95	343.	724.18	4.2279E-02
RU-103	1683.	497.08	1.5593E-02
RURH-106	720.	621.93	2.3308E-01
AG-108m	331.	722.94	9.9639E-03
CD-109	2233.	88.03	2.7140E-01
AG-110M	3556.	657.76	2.8227E-02
SN-113	2315.	391.69	2.0377E-02
TE-123m	2699.	159.00	8.2858E-03
SB-124	724.	602.73	1.1519E-02
SB-125	2615.	427.89	5.1563E-02
TE-125m	2344.	109.27	2.6136E+00
I-129	2170.	39.60	5.0184E+00
I-131	2292.	364.48	1.4904E-02
XE-131m	2704.	163.93	3.5876E-01
BA-133	2202.	356.02	1.8584E-02
CS-134	759.	604.70	1.1841E-02
CS-136	262.	818.51	9.0241E-03
CS-138	92.	1435.86	3.7821E-02
CE-139	2665.	165.85	8.7431E-03
BA-140	1106.	537.31	5.1136E-02
LA-140	90.	1596.21	1.0002E-02
CE-141	2575.	145.44	1.4021E-02
CE-144	2586.	133.51	6.1804E-02
CEPR-144	2586.	133.51	1.2349E-01
EU-152	2515.	121.78	2.4737E-02
EU-154	166.	1274.51	2.9841E-02
EU-155	2322.	86.54	3.1501E-02
HF-181	2079.	482.18	1.8981E-02
TA-182	2120.	67.75	4.6473E-02
HG-203	2574.	279.20	1.2016E-02
BI-207	803.	569.70	1.1490E-02

4550

DATA
NOT USED

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002598-SAM

Acquisition date : 18-NOV-1998 15:56:27

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uci/L)
TL-208	2545.	277.36	1.5317E-01
PB-210	2057.	46.50	2.8860E+00
BI-212	346.	727.18	1.3973E-01
PB-212	3113.	238.63	2.1423E-02
BI-214	1116.	609.31	3.1384E-02
PB-214	2637.	351.92	1.3810E-01
RA-224	2939.	240.99	2.3285E-01
RA-226	3253.	186.10	2.3036E-01
AC-228	507.	911.21	5.0646E-02
TH-228	2359.	84.37	8.6715E-01
TH-229	2228.	88.47	3.9249E-02
U-232	1994.	57.78	1.7542E+01
PA-233	2410.	312.17	2.7380E-02
UTH-233	2759.	245.34	8.1870E+00
PA-234M	239.	1001.03	1.5493E+00
TH-234	2122.	63.29	6.4727E-01
U-235	3252.	185.71	1.4004E-02
NP-237	2329.	86.48	8.3569E-02
U-237	1984.	59.54	8.9240E-02
NP-238	240.	984.45	3.6458E-02
NP-239	2357.	106.12	3.3768E-02
PU-239	2449.	129.30	1.0516E+02
AM-241	1984.	59.54	8.5295E-02
AM-243	2409.	74.67	2.5047E-02

DATA
NOT USED

*****
 * 222-S Laboratory Counting Room 18-NOV-1998 23:52:15.39 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27168
 Sample ID: S98T2598-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4387.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:21.04 sec
 Dead Time: 0.2%

Verified by:

R. Jones 11-19-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 18-NOV-1998 21:21:23.92
 Decayed to: 18-NOV-1998 21:21:23.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.53*	77848	1.63	1323.11	1314	18	0.8		CS-137	8.76

DATA
NOT USED

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2598-DUP

Acquisition date : 18-NOV-1998 21:21:23

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma
			Uncorrected	Decay Corr		
			uCi/L	uCi/L	2-Sigma Error	%Error
CS-137	30.00Y	1.00	8.758E+00	8.758E+00	0.067E+00	0.76
Total Activity :			8.758E+00	8.758E+00		

Grand Total Activity : 8.758E+00 8.758E+00

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

DATA
NOT USED

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2598-DUP

Acquisition date : 18-NOV-1998 21:21:23

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2961.	477.59	1.7412E-01
NA-24	103.	1368.55	9.1756E-03
K-40	2246.	1460.75	3.8427E-01
AR-41	141.	1293.60	1.4974E-02
SC-46	360.	1120.55	1.3436E-02
CR-51	2652.	320.08	1.1270E-01
MN-54	266.	834.83	9.1750E-03
CO-56	259.	846.76	9.1739E-03
CO-58	270.	810.78	9.0864E-03
FE-59	191.	1099.25	1.7259E-02
CO-60	144.	1332.50	9.9363E-03
ZN-65	214.	1115.55	2.0515E-02
SE-75	2977.	264.66	1.7097E-02
KR-85	1851.	514.00	3.5372E+00
SR-85	1849.	514.01	1.5991E-02
Y-88	37.	1836.06	7.1618E-03
Y-91	197.	1204.67	3.5506E+00
NB-94	243.	871.09	9.0984E-03
NB-95	308.	765.78	9.2049E-03
ZR-95	299.	756.73	1.6430E-02
ZRNB-95	322.	724.18	4.1030E-02
RU-103	2059.	497.08	1.7222E-02
RURH-106	847.	621.93	2.5238E-01
AG-108m	317.	722.94	9.7504E-03
CD-109	2592.	88.03	2.9220E-01
AG-110M	4317.	657.76	3.1075E-02
SN-113	2842.	391.69	2.2550E-02
TE-123m	3126.	159.00	8.9103E-03
SB-124	871.	602.73	1.2614E-02
SB-125	3008.	427.89	5.5252E-02
TE-125m	2647.	109.27	2.7752E+00
I-129	2451.	39.60	5.3298E+00
I-131	2727.	364.48	1.6242E-02
XE-131m	2986.	163.93	3.7685E-01
BA-133	2679.	356.02	2.0475E-02
CS-134	914.	604.70	1.2971E-02
CS-136	274.	818.51	9.2122E-03
CS-138	111.	1435.86	4.1408E-02
CE-139	2905.	165.85	9.1231E-03
BA-140	1284.	537.31	5.5026E-02
LA-140	74.	1596.21	9.0951E-03
CE-141	2919.	145.44	1.4917E-02
CE-144	2926.	133.51	6.5701E-02
CEPR-144	2926.	133.51	1.3128E-01
EU-152	2907.	121.78	2.6575E-02
EU-154	151.	1274.51	2.8508E-02
EU-155	2706.	86.54	3.3971E-02
HF-181	2474.	482.18	2.0683E-02
TA-182	2410.	67.75	4.9514E-02
HG-203	2925.	279.20	1.2798E-02
BI-207	942.	569.70	1.2426E-02

DATA
NOT USED

4554

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2598-DUP

Acquisition date : 18-NOV-1998 21:21:23

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2958.	277.36	1.6499E-01
PB-210	2351.	46.50	3.0828E+00
BI-212	398.	727.18	1.4950E-01
PB-212	3562.	238.63	2.2901E-02
BI-214	1281.	609.31	3.3592E-02
PB-214	3013.	351.92	1.4756E-01
RA-224	3318.	240.99	2.4725E-01
RA-226	3903.	186.10	2.5210E-01
AC-228	493.	911.21	4.9965E-02
TH-228	2692.	84.37	9.2568E-01
TH-229	2565.	88.47	4.2074E-02
U-232	2381.	57.78	1.9147E+01
PA-233	2772.	312.17	2.9342E-02
UTH-233	3222.	245.34	8.8396E+00
PA-234M	240.	1001.03	1.5542E+00
TH-234	2454.	63.29	6.9542E-01
U-235	3872.	185.71	1.5268E-02
NP-237	2709.	86.48	9.0059E-02
U-237	2420.	59.54	9.8433E-02
NP-238	218.	984.45	3.4815E-02
NP-239	2571.	106.12	3.5246E-02
PU-239	2900.	129.30	1.1433E+02
AM-241	2420.	59.54	9.4083E-02
AM-243	2769.	74.67	2.6835E-02

DATA
NOT USED

LABCORE Completed Worklist Report for Worklist# 27278

Analyst: scl

Instrument: GEA02

Book#: 28B57

Method: LA-548-121

Rev/Mod

F-φ

Worklist Comment: Rerun of WL 27143,BY-105 (2) mcb

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@GEA-01 C060-02 SOLID		4.501e-03	4.73e-03	105.088	% Recovery
1 STD	0		@GEA-01 C060-02E SOLID		1	3.22	3.220	% Ct Error
1 STD	0		@GEA-01 CS13702 SOLID		7.445e-03	7.79e-03	104.634	% Recovery
1 STD	0		@GEA-01 CS13702E SOLID		1	3.00	3.000	% Ct Error
2 BLNK-PREP	0		@GEA-01 C060-02 SOLID		1	8.32e-3		uCi/g
2 BLNK-PREP	0		@GEA-01 CS13702 SOLID		1	3.36e-02	0.034	uCi/g
3 SAMPLE	S98T002986	0 F	@GEA-01 C060-02 SOLID		N/A	9.630e-03	9.63e-003	uCi/g
3 SAMPLE	S98T002986	0 F	@GEA-01 C060-02E SOLID		N/A	n/a		% Ct. Error
3 SAMPLE	S98T002986	0 F	@GEA-01 CS13702 SOLID		N/A	7.605e+01		uCi/g
3 SAMPLE	S98T002986	0 F	@GEA-01 CS13702E SOLID		N/A	0.370		% Ct. Error
4 DUP	S98T002986	0 F	@GEA-01 C060-02 SOLID		<9.63e-3	<1.13e-2		RPD
4 DUP	S98T002986	0 F	@GEA-01 C060-02E SOLID		1	n/a		% Ct Error
4 DUP	S98T002986	0 F	@GEA-01 CS13702 SOLID		7.61e+01	4.86e+01	44.106	RPD
4 DUP	S98T002986	0 F	@GEA-01 CS13702E SOLID		1.00	0.460	0.460	% Ct Error

Comments Section:

Comments for sample# S98T002986 and test @GEA-01 .

DL=0 => n/a.

DL=0 => n/a.

Final page for worklist# 27278

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature Date 12-1-98

LABCORE Data Entry Template for Worklist# 27278

Analyst: S.L. Instrument: GEA00 2 Book# 28857

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: Rerun of WL 27143,BY-105 (2) mcb

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@GEA-01	SOLID		
2 BLNK-PREP			@GEA-01	SOLID		
3 SAMPLE	S98T002986 0 F		@GEA-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E					
4 DUP	S98T002986 0 F		@GEA-01	SOLID		

Final page for worklist # 27278

Sue La 11-30-98
Analyst Signature Date

M.B. 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 11:02:24.96 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27278
 Sample ID: WL27278-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2916.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:21.16 sec
 Dead Time: 0.7%

Verified by:

John Polyea, Dec 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 10:11:31.79
 Decayed to: 30-NOV-1998 10:11:31.79
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.61*	5353	1.47	1323.21	1315	16	3.0		CS-137	7.79
0	1173.15	2420	1.81	2346.34	2338	17	4.4		CO-60	4.80
0	1332.40	2111	1.89	2664.88	2655	18	4.7		CO-60	4.65
0	1460.48*	14	2.03	2921.09	2912	161	53.6		K-40	0.321

Summary of Nuclide Activity

Page : 2

Sample ID : WL27278-STD

Acquisition date : 30-NOV-1998 10:11:31

Total number of lines in spectrum	4	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	4	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~3.212E-01~~	~~3.212E-01~~	~~4.933E-01~~	~~153.59~~	~~070~~
CO-60	5.27Y	1.00	4.727E+00	4.727E+00	0.152E+00	3.22	
CS-137	30.00Y	1.00	7.786E+00	7.786E+00	0.233E+00	3.00	12/1/98
Total Activity :			1.283E+01	1.283E+01			

Grand Total Activity :	1.283E+01	1.283E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27278-STD

Acquisition date : 30-NOV-1998 10:11:31

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	267.	477.59	7.0268E-01
NA-24	12.	1368.55	4.2942E-02
AR-41	24.	1293.60	6.4639E-02
SC-46	122.	1120.55	1.0328E-01
CR-51	286.	320.08	5.1415E-01
MN-54	118.	834.83	8.0300E-02
CO-56	144.	846.76	8.9315E-02
CO-58	128.	810.78	8.1836E-02
FE-59	117.	1099.25	1.7648E-01
ZN-65	120.	1115.55	2.0124E-01
SE-75	362.	264.66	8.4225E-02
KR-85	198.	514.00	1.5546E+01
SR-85	198.	514.01	7.0299E-02
Y-88	7.	1836.06	4.5612E-02
Y-91	37.	1204.67	2.0871E+01
NB-94	138.	871.09	8.9361E-02
NB-95	124.	765.78	7.6699E-02
ZR-95	98.	756.73	1.2447E-01
ZRNB-95	100.	724.18	3.0256E-01
RU-103	203.	497.08	7.2652E-02
RURH-106	120.	621.93	1.2678E+00
AG-108m	109.	722.94	7.5482E-02
CD-109	287.	88.03	1.3650E+00
AG-110M	191.	657.76	8.7208E-02
SN-113	267.	391.69	9.4409E-02
TE-123m	322.	159.00	4.2127E-02
SB-124	134.	602.73	6.6125E-02
SB-125	229.	427.89	2.0743E-01
TE-125m	318.	109.27	1.3991E+01
I-129	273.	39.60	1.5347E+01
I-131	275.	364.48	7.0604E-02
XE-131m	311.	163.93	1.7859E+00
BA-133	241.	356.02	8.4734E-02
CS-134	157.	604.70	7.1477E-02
CS-136	106.	818.51	7.5346E-02
CS-138	5.	1435.86	6.5327E-02
CE-139	339.	165.85	4.5741E-02
BA-140	163.	537.31	2.6243E-01
LA-140	6.	1596.21	3.8552E-02
CE-141	338.	145.44	7.4731E-02
CE-144	318.	133.51	3.1937E-01
CEPR-144	318.	133.51	6.3818E-01
EU-152	335.	121.78	1.3233E-01
EU-154	19.	1274.51	1.4082E-01
EU-155	296.	86.54	1.5720E-01
HF-181	256.	482.18	8.9302E-02
TA-182	293.	67.75	2.2183E-01
HG-203	317.	279.20	5.9270E-02
BI-207	145.	569.70	6.5188E-02
TL-208	299.	277.36	7.3957E-01
PB-210	262.	46.50	1.0361E+01

4560

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27278-STD

Acquisition date : 30-NOV-1998 10:11:31

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	106.	727.18	1.0234E+00
PB-212	355.	238.63	1.0326E-01
BI-214	150.	609.31	1.5357E-01
PB-214	281.	351.92	2.7974E-01
RA-224	358.	240.99	1.1511E+00
RA-226	385.	186.10	1.1549E+00
AC-228	173.	911.21	3.8837E-01
TH-228	289.	84.37	4.2137E+00
TH-229	291.	88.47	1.9919E-01
U-232	265.	57.78	7.5588E+01
PA-233	304.	312.17	1.3522E-01
UTH-233	357.	245.34	4.1908E+01
PA-234M	143.	1001.03	1.5627E+01
TH-234	264.	63.29	2.8427E+00
U-235	391.	185.71	7.0741E-02
NP-237	298.	86.48	4.1778E-01
U-237	303.	59.54	4.1704E-01
NP-238	127.	984.45	3.4231E-01
NP-239	310.	106.12	1.7565E-01
PU-239	358.	129.30	5.9079E+02
AM-241	303.	59.54	4.0003E-01
AM-243	269.	74.67	1.1197E-01

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 13:45:22.46 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27278
 Sample ID: WL27278-BLNK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2917.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:27.11 sec
 Dead Time: 0.3%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 11:10:48.67
 Decayed to: 30-NOV-1998 11:10:48.67
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	34.60	202	1.58	69.27	64	11 50.1			
0	661.69*	137	1.45	1323.37	1318	11 40.3		CS-137	6.632E-02
0	969.24*	29	1.54	1938.49	1934	10 91.2			
0	1592.46	12	1.24	3185.09	3179	10127.6			

Summary of Nuclide Activity

Page : 2

Sample ID : WL27278-BLNK

Acquisition date : 30-NOV-1998 11:10:48

Total number of lines in spectrum	4	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	2	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma
			Uncorrected	Decay Corr		
			uCi/L	uCi/L		%Error
CS-137	30.00Y	1.00	6.632E-02	6.632E-02	2.670E-02	40.25
Total Activity :			6.632E-02	6.632E-02		

Grand Total Activity :	6.632E-02	6.632E-02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
 Sample ID : WL27278-BLNK

HNF-1663 REV. 0

Page : 3
 Acquisition date : 30-NOV-1998 11:10:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	92.	477.59	1.4125E-01
NA-24	18.	1368.55	1.7866E-02
K-40	310.	1460.75	6.3090E-01
AR-41	26.	1293.60	2.9275E-02
SC-46	49.	1120.55	2.2347E-02
CR-51	163.	320.08	1.3078E-01
MN-54	39.	834.83	1.5916E-02
CO-56	47.	846.76	1.7527E-02
CO-58	37.	810.78	1.5215E-02
FE-59	41.	1099.25	3.5863E-02
CO-60	18.	1332.50	1.6398E-02
ZN-65	33.	1115.55	3.6747E-02
SE-75	166.	264.66	1.9274E-02
KR-85	113.	514.00	3.9602E+00
SR-85	112.	514.01	1.7877E-02
Y-88	10.	1836.06	1.7391E-02
Y-91	31.	1204.67	6.3968E+00
NB-94	44.	871.09	1.7507E-02
NB-95	61.	765.78	1.8242E-02
ZR-95	34.	756.73	2.5240E-02
ZRNB-95	57.	724.18	7.7658E-02
RU-103	76.	497.08	1.5165E-02
RURH-106	61.	621.93	3.0892E-01
AG-108m	46.	722.94	1.6720E-02
CD-109	192.	88.03	3.7507E-01
AG-110M	63.	657.76	1.7194E-02
SN-113	116.	391.69	2.1132E-02
TE-123m	207.	159.00	1.1330E-02
SB-124	74.	602.73	1.6678E-02
SB-125	100.	427.89	4.6525E-02
TE-125m	189.	109.27	3.6290E+00
I-129	138.	39.60	3.6899E+00
I-131	108.	364.48	1.5109E-02
XE-131m	220.	163.93	5.0428E-01
BA-133	117.	356.02	1.9984E-02
CS-134	69.	604.70	1.6150E-02
CS-136	35.	818.51	1.5091E-02
CS-138	15.	1435.86	7.2631E-02
CE-139	204.	165.85	1.1938E-02
BA-140	72.	537.31	5.9584E-02
LA-140	8.	1596.21	1.4335E-02
CE-141	227.	145.44	2.0568E-02
CE-144	213.	133.51	8.7744E-02
CEPR-144	213.	133.51	1.7534E-01
EU-152	219.	121.78	3.5906E-02
EU-154	33.	1274.51	6.0775E-02
EU-155	180.	86.54	4.1225E-02
HF-181	93.	482.18	1.8336E-02
TA-182	196.	67.75	6.0895E-02
HG-203	173.	279.20	1.4777E-02
BI-207	64.	569.70	1.4793E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27278-BLNK

Acquisition date : 30-NOV-1998 11:10:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	182.	277.36	1.9426E-01
PB-210	155.	46.50	2.6895E+00
BI-212	64.	727.18	2.6957E-01
PB-212	234.	238.63	2.8139E-02
BI-214	132.	609.31	4.8169E-02
PB-214	172.	351.92	1.6369E-01
RA-224	221.	240.99	3.0572E-01
RA-226	238.	186.10	3.0471E-01
AC-228	95.	911.21	9.7493E-02
TH-228	178.	84.37	1.1139E+00
TH-229	194.	88.47	5.4683E-02
U-232	157.	57.78	1.9575E+01
PA-233	142.	312.17	3.1225E-02
UTH-233	165.	245.34	9.6349E+00
PA-234M	30.	1001.03	2.5010E+00
TH-234	169.	63.29	7.6359E-01
U-235	239.	185.71	1.8591E-02
NP-237	179.	86.48	1.0896E-01
U-237	168.	59.54	1.0504E-01
NP-238	35.	984.45	6.3346E-02
NP-239	195.	106.12	4.7337E-02
PU-239	210.	129.30	1.5205E+02
AM-241	168.	59.54	1.0040E-01
AM-243	221.	74.67	3.3989E-02

*****

* 222-S Laboratory Counting Room 30-NOV-1998 16:33:50.77 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27278
 Sample ID: S98T2986-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2919.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:32:16.74 sec
 Dead Time: 1.5%

Verified by:

12 Dec 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 30-NOV-1998 14:01:03.23
 Decayed to: 30-NOV-1998 14:01:03.23
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.57*	309331	1.47	1323.14	1315	16	0.4		CS-137	150.

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2986-SAM

Acquisition date : 30-NOV-1998 14:01:03

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.500E+02	1.500E+02	0.006E+02	0.37	
Total Activity :			1.500E+02	1.500E+02			

Grand Total Activity :	1.500E+02	1.500E+02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2986-SAM

Acquisition date : 30-NOV-1998 14:01:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	9288.	477.59	1.3434E+00
NA-24	20.	1368.55	1.8642E-02
K-40	285.	1460.75	6.0577E-01
AR-41	23.	1293.60	2.7999E-02
SC-46	67.	1120.55	2.5970E-02
CR-51	7722.	320.08	8.6701E-01
MN-54	43.	834.83	1.6633E-02
CO-56	32.	846.76	1.4722E-02
CO-58	50.	810.78	1.7490E-02
FE-59	33.	1099.25	3.2754E-02
CO-60	25.	1332.50	1.8990E-02
ZN-65	33.	1115.55	3.6949E-02
SE-75	8824.	264.66	1.3531E-01
KR-85	4785.	514.00	2.4651E+01
SR-85	4784.	514.01	1.1150E-01
Y-88	7.	1836.06	1.5158E-02
Y-91	36.	1204.67	6.8878E+00
NB-94	43.	871.09	1.7228E-02
NB-95	59.	765.78	1.8050E-02
ZR-95	54.	756.73	3.1299E-02
ZRNB-95	70.	724.18	8.4936E-02
RU-103	5707.	497.08	1.2453E-01
RURH-106	1973.	621.93	1.6516E+00
AG-108m	66.	722.94	1.9921E-02
CD-109	7511.	88.03	2.2665E+00
AG-110M	6716.	657.76	1.6677E-01
SN-113	8462.	391.69	1.7215E-01
TE-123m	8845.	159.00	7.1691E-02
SB-124	2008.	602.73	8.2238E-02
SB-125	9382.	427.89	4.2858E-01
TE-125m	8096.	109.27	2.2946E+01
I-129	7164.	39.60	2.5478E+01
I-131	7756.	364.48	1.2198E-01
XE-131m	8817.	163.93	3.0931E+00
BA-133	7644.	356.02	1.5433E-01
CS-134	1968.	604.70	8.1779E-02
CS-136	39.	818.51	1.5768E-02
CS-138	16.	1435.86	7.4217E-02
CE-139	8642.	165.85	7.5120E-02
BA-140	3491.	537.31	3.9187E-01
LA-140	8.	1596.21	1.4679E-02
CE-141	8789.	145.44	1.2401E-01
CE-144	8901.	133.51	5.4868E-01
CEPR-144	8900.	133.51	1.0964E+00
EU-152	8555.	121.78	2.1752E-01
EU-154	20.	1274.51	4.8050E-02
EU-155	7460.	86.54	2.5605E-01
HF-181	7552.	482.18	1.5727E-01
TA-182	6994.	67.75	3.5178E-01
HG-203	8396.	279.20	9.9141E-02
BI-207	2309.	569.70	8.3746E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2986-SAM

Acquisition date : 30-NOV-1998 14:01:03

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	8444.	277.36	1.2755E+00
PB-210	7055.	46.50	1.7468E+01
BI-212	72.	727.18	2.8372E-01
PB-212	9919.	238.63	1.7757E-01
BI-214	2091.	609.31	1.8479E-01
PB-214	7845.	351.92	1.0767E+00
RA-224	9588.	240.99	1.9509E+00
RA-226	11197.	186.10	2.0265E+00
AC-228	63.	911.21	8.0418E-02
TH-228	7386.	84.37	6.9145E+00
TH-229	7508.	88.47	3.2842E-01
U-232	7024.	57.78	1.2614E+02
PA-233	7703.	312.17	2.2108E-01
UTH-233	9411.	245.34	6.9988E+01
PA-234M	28.	1001.03	2.4292E+00
TH-234	6937.	63.29	4.7205E+00
U-235	11129.	185.71	1.2287E-01
NP-237	7467.	86.48	6.7853E-01
U-237	7135.	59.54	6.5992E-01
NP-238	41.	984.45	6.8350E-02
NP-239	7864.	106.12	2.9046E-01
PU-239	8738.	129.30	9.4947E+02
AM-241	7135.	59.54	6.3072E-01
AM-243	7129.	74.67	1.8690E-01

*****

* 222-S Laboratory Counting Room 30-NOV-1998 19:18:54.10 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27278
 Sample ID: S98T2986-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Newright

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA2
 File Number: dka300:[spec.GEA2]2g2920.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:31:50.22 sec
 Dead Time: 1.2%

Verified by:

John Relyea Dec 28

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 30-NOV-1998 16:46:32.85
 Decayed to: 30-NOV-1998 16:46:32.85
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA2]2GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 19:57:04.05
 Date of last efficiency calibration: 10-NOV-1997 10:04:32.95

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.57*	198877	1.47	1323.14	1315	16	0.5		CS-137	96.4
0	1476.44	11	1.20	2953.00	2950	7	60.3			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2986-DUP

Acquisition date : 30-NOV-1998 16:46:32

Total number of lines in spectrum	2	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	1	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	9.643E+01	9.643E+01	0.044E+01	0.46	
Total Activity :			9.643E+01	9.643E+01			

Grand Total Activity :	9.643E+01	9.643E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2986-DUP

Acquisition date : 30-NOV-1998 16:46:32

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	6072.	477.59	1.0878E+00
NA-24	23.	1368.55	1.9842E-02
K-40	308.	1460.75	6.2822E-01
AR-41	30.	1293.60	3.1840E-02
SC-46	63.	1120.55	2.5224E-02
CR-51	5062.	320.08	7.0306E-01
MN-54	40.	834.83	1.6174E-02
CO-56	56.	846.76	1.9095E-02
CO-58	52.	810.78	1.7848E-02
FE-59	38.	1099.25	3.4960E-02
CO-60	36.	1332.50	2.2473E-02
ZN-65	38.	1115.55	3.9177E-02
SE-75	5719.	264.66	1.0909E-01
KR-85	3098.	514.00	1.9877E+01
SR-85	3097.	514.01	8.9900E-02
Y-88	7.	1836.06	1.5158E-02
Y-91	31.	1204.67	6.3969E+00
NB-94	35.	871.09	1.5774E-02
NB-95	58.	765.78	1.7823E-02
ZR-95	62.	756.73	3.3296E-02
ZRNB-95	48.	724.18	7.1290E-02
RU-103	3761.	497.08	1.0127E-01
RURH-106	1241.	621.93	1.3145E+00
AG-108m	51.	722.94	1.7601E-02
CD-109	4988.	88.03	1.8499E+00
AG-110M	4254.	657.76	1.3297E-01
SN-113	5467.	391.69	1.3858E-01
TE-123m	5733.	159.00	5.7803E-02
SB-124	1296.	602.73	6.6282E-02
SB-125	6305.	427.89	3.5180E-01
TE-125m	5181.	109.27	1.8386E+01
I-129	4575.	39.60	2.0396E+01
I-131	5130.	364.48	9.9345E-02
XE-131m	5723.	163.93	2.4958E+00
BA-133	5021.	356.02	1.2527E-01
CS-134	1299.	604.70	6.6635E-02
CS-136	47.	818.51	1.7251E-02
CS-138	17.	1435.86	7.5390E-02
CE-139	5678.	165.85	6.0981E-02
BA-140	2222.	537.31	3.1342E-01
LA-140	7.	1596.21	1.3670E-02
CE-141	5786.	145.44	1.0076E-01
CE-144	5695.	133.51	4.3955E-01
CEPR-144	5694.	133.51	8.7823E-01
EU-152	5657.	121.78	1.7714E-01
EU-154	21.	1274.51	4.9358E-02
EU-155	5023.	86.54	2.1042E-01
HF-181	4707.	482.18	1.2437E-01
TA-182	4577.	67.75	2.8504E-01
HG-203	5435.	279.20	7.9892E-02
BI-207	1508.	569.70	6.7875E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2986-DUP

Acquisition date : 30-NOV-1998 16:46:32

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	5425.	277.36	1.0240E+00
PB-210	4526.	46.50	1.4015E+01
BI-212	65.	727.18	2.6990E-01
PB-212	6575.	238.63	1.4476E-01
BI-214	1349.	609.31	1.4889E-01
PB-214	5151.	351.92	8.7148E-01
RA-224	6436.	240.99	1.6003E+00
RA-226	7451.	186.10	1.6552E+00
AC-228	75.	911.21	8.6815E-02
TH-228	4968.	84.37	5.6793E+00
TH-229	5002.	88.47	2.6846E-01
U-232	4588.	57.78	1.0212E+02
PA-233	5109.	312.17	1.8033E-01
UTH-233	6280.	245.34	5.7248E+01
PA-234M	41.	1001.03	2.9127E+00
TH-234	4550.	63.29	3.8292E+00
U-235	7364.	185.71	1.0007E-01
NP-237	5032.	86.48	5.5783E-01
U-237	4595.	59.54	5.3046E-01
NP-238	38.	984.45	6.5445E-02
NP-239	5170.	106.12	2.3587E-01
PU-239	5830.	129.30	7.7662E+02
AM-241	4595.	59.54	5.0699E-01
AM-243	4734.	74.67	1.5255E-01

LABCORE Completed Worklist Report for Worklist# 27288

Analyst: gl

Instrument: GEA04

Book#: 28B57

Method: LA-548-121 Rev/Mod F-φ

Worklist Comment: BY-105

@GEA-01

RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@GEA-01 C060-02 SOLID		4.501e-03	4.74e-03	105.310 % Recovery	
1 STD	0		@GEA-01 C060-02E SOLID		1	1.57	1.570 % Ct Error	
1 STD	0		@GEA-01 CS13702 SOLID		7.445e-03	7.88e-03	105.843 % Recovery	
1 STD	0		@GEA-01 CS13702E SOLID		1	1.47	1.470 % Ct Error	
2 BLNK-PREP	0		@GEA-01 C060-02 SOLID		1	<4.85e-3	uCi/g	
2 BLNK-PREP	0		@GEA-01 CS13702 SOLID		1	<7.58e-3	uCi/g	
3 SAMPLE	S98T002598	0 F	@GEA-01 C060-02 SOLID		N/A	< 5.21e-03	5.21e-003 uCi/g	
3 SAMPLE	S98T002598	0 F	@GEA-01 C060-02E SOLID		N/A	n/a	% Ct. Error	
3 SAMPLE	S98T002598	0 F	@GEA-01 CS13702 SOLID		N/A	3.431e+00	uCi/g	
3 SAMPLE	S98T002598	0 F	@GEA-01 CS13702E SOLID		N/A	0.850	% Ct. Error	
4 DUP	S98T002598	0 F	@GEA-01 C060-02 SOLID		<5.21e-3	<4.83e-3	RPD	
4 DUP	S98T002598	0 F	@GEA-01 C060-02E SOLID		1	n/a	% Ct Error	
4 DUP	S98T002598	0 F	@GEA-01 CS13702 SOLID		3.43e+00	4.39e+00	24.552 RPD	
4 DUP	S98T002598	0 F	@GEA-01 CS13702E SOLID		1.00	0.750	0.750 % Ct Error	

Comments Section:

Comments for sample# S98T002598 and test @GEA-01 .

DL=0 => n/a.
DL=0 => n/a.
DL=0 => n/a.
DL=0 => n/a.

Final page for worklist# 27288

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

RA Jones 12-1-98
Reviewer Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27288

Analyst: gl Instrument: GEA00 _____ Book# 28857

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY-105 @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T002598	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T002598	0	F	@GEA-01	SOLID		

Final page for worklist # 27288

Gay L Platt 11-30-98
Analyst Signature Date

MeB 12/1/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 15:39:45.94 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27288
 Sample ID: WL27288-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4446.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:09.13 sec
 Dead Time: 0.3%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 14:49:01.87
 Decayed to: 30-NOV-1998 14:49:01.87
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	59.41	256	1.30	118.97	115	9	57.2		AM-241	0.313
0	661.50*	23352	1.66	1323.07	1315	17	1.5		CS-137	7.88
0	1173.02	10239	1.94	2346.15	2336	19	2.2		CO-60	4.71
0	1332.33	9348	2.07	2664.80	2654	22	2.2		CO-60	4.77

Summary of Nuclide Activity
Sample ID : WL27288-STD

Page : 2

Acquisition date : 30-NOV-1998 14:49:01

Total number of lines in spectrum	4	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	4	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.737E+00	4.737E+00	0.074E+00	1.57	
CS-137	30.00Y	1.00	7.882E+00	7.882E+00	0.116E+00	1.47	
~~AM-241~~	~~432.20Y*****~~		~~3.125E-01~~	~~3.125E-01~~	~~1.789E-01~~	~~57.25~~	~~12-14~~
Total Activity :			1.293E+01	1.293E+01			

Grand Total Activity : 1.293E+01 1.293E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27288-STD

Acquisition date : 30-NOV-1998 14:49:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1300.	477.59	3.4773E-01
NA-24	47.	1368.55	1.8366E-02
K-40	778.	1460.75	6.8426E-01
AR-41	108.	1293.60	2.9873E-02
SC-46	557.	1120.55	4.9823E-02
CR-51	1277.	320.08	2.3555E-01
MN-54	570.	834.83	3.9865E-02
CO-56	617.	846.76	4.1951E-02
CO-58	596.	810.78	4.0008E-02
FE-59	554.	1099.25	8.6665E-02
ZN-65	519.	1115.55	9.4638E-02
SE-75	1419.	264.66	3.5579E-02
KR-85	907.	514.00	7.4699E+00
SR-85	906.	514.01	3.3762E-02
Y-88	20.	1836.06	1.6423E-02
Y-91	209.	1204.67	1.0933E+01
NB-94	634.	871.09	4.3523E-02
NB-95	531.	765.78	3.5939E-02
ZR-95	534.	756.73	6.5341E-02
ZRNB-95	493.	724.18	1.5126E-01
RU-103	924.	497.08	3.4803E-02
RURH-106	626.	621.93	6.5286E-01
AG-108m	515.	722.94	3.7033E-02
CD-109	1147.	88.03	5.8644E-01
AG-110M	1661.	657.76	5.8136E-02
SN-113	1172.	391.69	4.3699E-02
TE-123m	1275.	159.00	1.7166E-02
SB-124	612.	602.73	3.1830E-02
SB-125	1250.	427.89	1.0748E-01
TE-125m	1201.	109.27	5.6384E+00
I-129	1099.	39.60	1.0770E+01
I-131	1170.	364.48	3.2000E-02
XE-131m	1302.	163.93	7.4906E-01
BA-133	1145.	356.02	4.0399E-02
CS-134	629.	604.70	3.2407E-02
CS-136	588.	818.51	3.9978E-02
CS-138	33.	1435.86	3.4459E-02
CE-139	1334.	165.85	1.8637E-02
BA-140	733.	537.31	1.2517E-01
LA-140	32.	1596.21	1.8183E-02
CE-141	1320.	145.44	3.0227E-02
CE-144	1254.	133.51	1.2975E-01
CEPR-144	1254.	133.51	2.5925E-01
EU-152	1292.	121.78	5.3427E-02
EU-154	137.	1274.51	8.1513E-02
EU-155	1169.	86.54	6.7377E-02
HF-181	1041.	482.18	4.0474E-02
TA-182	1089.	67.75	1.0042E-01
HG-203	1307.	279.20	2.5791E-02
BI-207	659.	569.70	3.1297E-02
TL-208	1313.	277.36	3.3153E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27288-STD

Acquisition date : 30-NOV-1998 14:49:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1092.	46.50	6.3192E+00
BI-212	491.	727.18	4.9680E-01
PB-212	1602.	238.63	4.6293E-02
BI-214	823.	609.31	8.1075E-02
PB-214	1325.	351.92	1.3271E-01
RA-224	1569.	240.99	5.0899E-01
RA-226	1583.	186.10	4.8417E-01
AC-228	776.	911.21	1.8702E-01
TH-228	1271.	84.37	1.9176E+00
TH-229	1132.	88.47	8.4327E-02
U-232	1220.	57.78	4.1307E+01
PA-233	1291.	312.17	6.0319E-02
UTH-233	1470.	245.34	1.8000E+01
PA-234M	582.	1001.03	7.1655E+00
TH-234	1068.	63.29	1.3844E+00
U-235	1586.	185.71	2.9470E-02
NP-237	1171.	86.48	1.7863E-01
U-237	1280.	59.54	2.1496E-01
NP-238	624.	984.45	1.7203E-01
NP-239	1153.	106.12	7.0501E-02
PU-239	1262.	129.30	2.2750E+02
AM-243	1228.	74.67	5.3900E-02

*****

* 222-S Laboratory Counting Room 30-NOV-1998 18:28:02.77 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27288
Sample ID: WL27288-BLNK
Sample Size: 1.00000E-03 L
Dilution Factor: 1.00000E+00

Removed by:

Newright

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA4
File Number: dka300:[spec.GEA4]4g4447.cnf
Geometry: 41
Count Time: 0 02:30:00.00 sec
Real Time: 0 02:30:06.71 sec
Dead Time: 0.1%

Verified by:

John F. Lyle 12 Dec 98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 30-NOV-1998 15:57:21.74
Decayed to: 30-NOV-1998 15:57:21.74
Standard Deviations: 2
Analysis Library: ENVGEA
Analyst: MB
Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

***** No peaks found *****

Summary of Nuclide Activity

Page : 2

Sample ID : WL27288-BLNK

Acquisition date : 30-NOV-1998 15:57:21

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27288-BLNK

Acquisition date : 30-NOV-1998 15:57:21

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	548.	477.59	7.5961E-02
NA-24	102.	1368.55	9.1545E-03
K-40	2204.	1460.75	3.8075E-01
AR-41	138.	1293.60	1.4834E-02
SC-46	341.	1120.55	1.3092E-02
CR-51	793.	320.08	6.2186E-02
MN-54	302.	834.83	9.7617E-03
CO-56	223.	846.76	8.5463E-03
CO-58	244.	810.78	8.6417E-03
FE-59	219.	1099.25	1.8447E-02
CO-60	142.	1332.50	9.8830E-03
ZN-65	213.	1115.55	2.0487E-02
SE-75	966.	264.66	9.8186E-03
KR-85	714.	514.00	2.2141E+00
SR-85	711.	514.01	9.9996E-03
Y-88	54.	1836.06	8.5759E-03
Y-91	217.	1204.67	3.7137E+00
NB-94	260.	871.09	9.3995E-03
NB-95	311.	765.78	9.2423E-03
ZR-95	281.	756.73	1.5957E-02
ZRNB-95	319.	724.18	4.0806E-02
RU-103	466.	497.08	8.3060E-03
RURH-106	367.	621.93	1.6782E-01
AG-108m	282.	722.94	9.2217E-03
CD-109	951.	88.03	1.7831E-01
AG-110M	380.	657.76	9.4148E-03
SN-113	649.	391.69	1.0901E-02
TE-123m	1091.	159.00	5.3012E-03
SB-124	395.	602.73	8.5691E-03
SB-125	558.	427.89	2.4119E-02
TE-125m	966.	109.27	1.6889E+00
I-129	792.	39.60	3.0553E+00
I-131	704.	364.48	8.3387E-03
XE-131m	1096.	163.93	2.2983E-01
BA-133	704.	356.02	1.0607E-02
CS-134	410.	604.70	8.7729E-03
CS-136	269.	818.51	9.1301E-03
CS-137	837.	661.66	1.5437E-02
CS-138	109.	1435.86	4.0904E-02
CE-139	1083.	165.85	5.6095E-03
BA-140	411.	537.31	3.1528E-02
LA-140	80.	1596.21	9.4287E-03
CE-141	1088.	145.44	9.1702E-03
CE-144	1032.	133.51	3.9307E-02
CEPR-144	1033.	133.51	7.8550E-02
EU-152	1043.	121.78	1.6030E-02
EU-154	133.	1274.51	2.6805E-02
EU-155	1067.	86.54	2.1472E-02
HF-181	486.	482.18	9.3028E-03
TA-182	870.	67.75	2.9987E-02
HG-203	906.	279.20	7.1834E-03

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27288-BLNK

Acquisition date : 30-NOV-1998 15:57:21

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	399.	569.70	8.1635E-03
TL-208	1003.	277.36	9.6823E-02
PB-210	794.	46.50	1.8067E+00
BI-212	383.	727.18	1.4680E-01
PB-212	1313.	238.63	1.3989E-02
BI-214	877.	609.31	2.7879E-02
PB-214	1131.	351.92	9.0867E-02
RA-224	1123.	240.99	1.4487E-01
RA-226	1234.	186.10	1.4278E-01
AC-228	520.	911.21	5.1277E-02
TH-228	1139.	84.37	6.0570E-01
TH-229	915.	88.47	2.5325E-02
U-232	804.	57.78	1.1222E+01
PA-233	829.	312.17	1.6193E-02
UTH-233	1014.	245.34	4.9996E+00
PA-234M	235.	1001.03	1.5388E+00
TH-234	858.	63.29	4.1452E-01
U-235	1237.	185.71	8.6931E-03
NP-237	1078.	86.48	5.7178E-02
U-237	767.	59.54	5.5931E-02
NP-238	218.	984.45	3.4810E-02
NP-239	907.	106.12	2.1102E-02
PU-239	1057.	129.30	6.9523E+01
AM-241	767.	59.54	5.3460E-02
AM-243	1086.	74.67	1.6914E-02

*****
 * 222-S Laboratory Counting Room 30-NOV-1998 23:28:40.40 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27288
 Sample ID: S98T002598-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Reusing LT

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4448.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:14.80 sec
 Dead Time: 0.2%

Verified by:

John Delaney / Dec 88

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 30-NOV-1998 20:57:56.39
 Decayed to: 30-NOV-1998 20:57:56.39
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	609.02*	243	1.74	1218.10	1210	17	69.2			
0	661.50*	62170	1.66	1323.06	1314	18	0.9		CS-137	6.99
0	1730.88*	55	3.44	3462.04	3450	30	95.0			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002598-SAM

Acquisition date : 30-NOV-1998 20:57:56

Total number of lines in spectrum	3	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	2	66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	6.994E+00	6.994E+00	0.060E+00	0.85	
Total Activity :			6.994E+00	6.994E+00			

Grand Total Activity :	6.994E+00	6.994E+00

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002598-SAM

Acquisition date : 30-NOV-1998 20:57:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2454.	477.59	1.5866E-01
NA-24	120.	1368.55	9.8563E-03
K-40	2227.	1460.75	3.8267E-01
AR-41	153.	1293.60	1.5582E-02
SC-46	361.	1120.55	1.3461E-02
CR-51	2296.	320.08	1.0496E-01
MN-54	268.	834.83	9.2182E-03
CO-56	244.	846.76	8.9228E-03
CO-58	259.	810.78	8.8983E-03
FE-59	205.	1099.25	1.7864E-02
CO-60	165.	1332.50	1.0617E-02
ZN-65	203.	1115.55	2.0016E-02
SE-75	2570.	264.66	1.5898E-02
KR-85	1652.	514.00	3.3439E+00
SR-85	1649.	514.01	1.5116E-02
Y-88	39.	1836.06	7.3687E-03
Y-91	226.	1204.67	3.7852E+00
NB-94	256.	871.09	9.3291E-03
NB-95	278.	765.78	8.7528E-03
ZR-95	289.	756.73	1.6169E-02
ZRNB-95	310.	724.18	4.0259E-02
RU-103	1785.	497.08	1.6048E-02
RURH-106	717.	621.93	2.3259E-01
AG-108m	304.	722.94	9.5564E-03
CD-109	2364.	88.03	2.7919E-01
AG-110M	3559.	657.76	2.8242E-02
SN-113	2377.	391.69	2.0646E-02
TE-123m	2654.	159.00	8.2161E-03
SB-124	805.	602.73	1.2130E-02
SB-125	2552.	427.89	5.0944E-02
TE-125m	2335.	109.27	2.6086E+00
I-129	2169.	39.60	5.0169E+00
I-131	2225.	364.48	1.4687E-02
XE-131m	2714.	163.93	3.5942E-01
BA-133	2276.	356.02	1.8889E-02
CS-134	792.	604.70	1.2088E-02
CS-136	274.	818.51	9.2076E-03
CS-138	103.	1435.86	3.9777E-02
CE-139	2711.	165.85	8.8173E-03
BA-140	1092.	537.31	5.0817E-02
LA-140	70.	1596.21	8.8622E-03
CE-141	2676.	145.44	1.4290E-02
CE-144	2532.	133.51	6.1158E-02
CEPR-144	2531.	133.51	1.2220E-01
EU-152	2394.	121.78	2.4142E-02
EU-154	146.	1274.51	2.8054E-02
EU-155	2334.	86.54	3.1578E-02
HF-181	2122.	482.18	1.9170E-02
TA-182	2116.	67.75	4.6434E-02
HG-203	2502.	279.20	1.1847E-02
BI-207	896.	569.70	1.2128E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002598-SAM

Acquisition date : 30-NOV-1998 20:57:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2583.	277.36	1.5429E-01
PB-210	2015.	46.50	2.8565E+00
BI-212	384.	727.18	1.4688E-01
PB-212	3123.	238.63	2.1457E-02
BI-214	1264.	609.31	3.3364E-02
PB-214	2636.	351.92	1.3803E-01
RA-224	2939.	240.99	2.3282E-01
RA-226	3346.	186.10	2.3359E-01
AC-228	469.	911.21	4.8737E-02
TH-228	2338.	84.37	8.6339E-01
TH-229	2377.	88.47	4.0522E-02
U-232	2100.	57.78	1.7994E+01
PA-233	2365.	312.17	2.7125E-02
UTH-233	2689.	245.34	8.0830E+00
PA-234M	240.	1001.03	1.5538E+00
TH-234	2076.	63.29	6.4030E-01
U-235	3312.	185.71	1.4132E-02
NP-237	2336.	86.48	8.3692E-02
U-237	2061.	59.54	9.0924E-02
NP-238	210.	984.45	3.4177E-02
NP-239	2252.	106.12	3.3011E-02
PU-239	2521.	129.30	1.0667E+02
AM-241	2061.	59.54	8.6904E-02
AM-243	2380.	74.67	2.4901E-02

*****

* 222-S Laboratory Counting Room 1-DEC-1998 06:37:40.08 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27288
 Sample ID: S98T002598-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA4
 File Number: dka300:[spec.GEA4]4g4451.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:17.28 sec
 Dead Time: 0.2%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 1-DEC-1998 04:06:52.09
 Decayed to: 1-DEC-1998 04:06:52.09
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: NEW
 Background Subtract: DKA300:[SPEC.GEA4]4GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 10-SEP-1998 09:10:53.69
 Date of last efficiency calibration: 29-OCT-1997 14:44:52.60

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	511.14*	251	3.16	1022.34	1015	15	98.5			
0	661.50*	79511	1.65	1323.06	1314	18	0.8		CS-137	8.95
0	1119.85*	81	1.37	2239.80	2232	161	03.4			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T002598-DUP

Acquisition date : 1-DEC-1998 04:06:52

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	Flags
			Uncorrected	Decay Corr			
			uCi/L	uCi/L	2-Sigma Error	%Error	
CS-137	30.00Y	1.00	8.945E+00	8.945E+00	0.067E+00	0.75	
Total Activity :			8.945E+00	8.945E+00			

Grand Total Activity :	8.945E+00	8.945E+00

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T002598-DUP

Acquisition date : 1-DEC-1998 04:06:52

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	3026.	477.59	1.7598E-01
NA-24	104.	1368.55	9.2264E-03
K-40	2275.	1460.75	3.8673E-01
AR-41	116.	1293.60	1.3663E-02
SC-46	402.	1120.55	1.4178E-02
CR-51	2760.	320.08	1.1496E-01
MN-54	281.	834.83	9.4352E-03
CO-56	254.	846.76	9.0947E-03
CO-58	258.	810.78	8.8795E-03
FE-59	195.	1099.25	1.7445E-02
CO-60	141.	1332.50	9.8318E-03
ZN-65	223.	1115.55	2.0928E-02
SE-75	2996.	264.66	1.7151E-02
KR-85	1929.	514.00	3.6101E+00
SR-85	1927.	514.01	1.6319E-02
Y-88	36.	1836.06	7.0903E-03
Y-91	215.	1204.67	3.6967E+00
NB-94	239.	871.09	9.0246E-03
NB-95	309.	765.78	9.2127E-03
ZR-95	289.	756.73	1.6153E-02
ZRNB-95	320.	724.18	4.0890E-02
RU-103	1967.	497.08	1.6837E-02
RURH-106	838.	621.93	2.5095E-01
AG-108m	305.	722.94	9.5779E-03
CD-109	2748.	88.03	3.0072E-01
AG-110M	4573.	657.76	3.1975E-02
SN-113	2934.	391.69	2.2911E-02
TE-123m	3001.	159.00	8.7316E-03
SB-124	888.	602.73	1.2728E-02
SB-125	3137.	427.89	5.6417E-02
TE-125m	2776.	109.27	2.8413E+00
I-129	2475.	39.60	5.3546E+00
I-131	2769.	364.48	1.6365E-02
XE-131m	3137.	163.93	3.8611E-01
BA-133	2618.	356.02	2.0242E-02
CS-134	890.	604.70	1.2803E-02
CS-136	239.	818.51	8.6315E-03
CS-138	94.	1435.86	3.8223E-02
CE-139	3064.	165.85	9.3665E-03
BA-140	1291.	537.31	5.5185E-02
LA-140	65.	1596.21	8.5708E-03
CE-141	3010.	145.44	1.5146E-02
CE-144	3015.	133.51	6.6682E-02
CEPR-144	3015.	133.51	1.3324E-01
EU-152	2883.	121.78	2.6467E-02
EU-154	154.	1274.51	2.8729E-02
EU-155	2774.	86.54	3.4396E-02
HF-181	2582.	482.18	2.1123E-02
TA-182	2595.	67.75	5.1357E-02
HG-203	2966.	279.20	1.2888E-02
BI-207	938.	569.70	1.2404E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T002598-DUP

Acquisition date : 1-DEC-1998 04:06:52

Nuclide	Bkgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	3012.	277.36	1.6648E-01
PB-210	2378.	46.50	3.1003E+00
BI-212	366.	727.18	1.4343E-01
PB-212	3510.	238.63	2.2735E-02
BI-214	1296.	609.31	3.3779E-02
PB-214	3085.	351.92	1.4922E-01
RA-224	3413.	240.99	2.5072E-01
RA-226	3929.	186.10	2.5292E-01
AC-228	487.	911.21	4.9662E-02
TH-228	2700.	84.37	9.2704E-01
TH-229	2724.	88.47	4.3346E-02
U-232	2436.	57.78	1.9364E+01
PA-233	2865.	312.17	2.9819E-02
UTH-233	3219.	245.34	8.8355E+00
PA-234M	240.	1001.03	1.5539E+00
TH-234	2490.	63.29	7.0054E-01
U-235	3892.	185.71	1.5307E-02
NP-237	2782.	86.48	9.1241E-02
U-237	2475.	59.54	9.9536E-02
NP-238	197.	984.45	3.3143E-02
NP-239	2710.	106.12	3.6179E-02
PU-239	3052.	129.30	1.1726E+02
AM-241	2475.	59.54	9.5137E-02
AM-243	2773.	74.67	2.6853E-02

LABCORE Completed Worklist Report for Worklist# 27626

Analyst: crj

Instrument: GEA03

Book#: 28B57

Method: LA-548-121

Rev/Mod F-φ

Worklist Comment: BY-105 @GEA-01 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@GEA-01 C060-02	LIQUID	4.465e-03	4.77e-03	106.931	% Recovery
1 STD	0		@GEA-01 C060-02R	LIQUID	1	1.65	1.650	% Ct Error
1 STD	0		@GEA-01 CS13702	LIQUID	7.434e-03	7.51e-03	101.022	% Recovery
1 STD	0		@GEA-01 CS13702R	LIQUID	1	1.51	1.510	% Ct Error
2 BLNK	0		@GEA-01 C060-02	LIQUID	1	<7.93e-4		uCi/mL
2 BLNK	0		@GEA-01 CS13702	LIQUID	1	<1.65e-3		uCi/mL
3 SAMPLE	S98T003384	0	@GEA-01 C060-02	LIQUID	N/A	< 3.729e-03	3.73e-003	uCi/mL
3 SAMPLE	S98T003384	0	@GEA-01 C060-02R	LIQUID	N/A	n/a		% Ct. Error
3 SAMPLE	S98T003384	0	@GEA-01 CS13702	LIQUID	N/A	1.960e+02		uCi/mL
3 SAMPLE	S98T003384	0	@GEA-01 CS13702R	LIQUID	N/A	0.120		% Ct. Error
4 DUP	S98T003384	0	@GEA-01 C060-02	LIQUID	<3.73e-3	<3.93e-3		RPD
4 DUP	S98T003384	0	@GEA-01 C060-02R	LIQUID	1	n/a		% Ct Error
4 DUP	S98T003384	0	@GEA-01 CS13702	LIQUID	1.96e+02	2.01e+02	2.519	RPD
4 DUP	S98T003384	0	@GEA-01 CS13702R	LIQUID	1	0.120	0.120	% Ct Error

Comments Section:

Comments for sample# S98T003384 and test @GEA-01 .

DL=0 => n/a.

Final page for worklist# 27626

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature

12-28-98
Date

LABCORE Data Entry Template for Worklist# 27626

Analyst: _____ Instrument: GEA00 _____ Book# _____

Method: LA-548-121 Rev/Mod _____

Worklist Comment: BY-105 @GEA-01 RLC

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	LIQUID		
2	BLNK				@GEA-01	LIQUID		
3	SAMPLE	S98T003384 0			@GEA-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E								
4	DUP	S98T003384 0			@GEA-01	LIQUID		

Final page for worklist # 27626

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

*****
 * 222-S Laboratory Counting Room 22-DEC-1998 19:18:25.45 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27626
 Sample ID: WL27626-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Em Brice

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3038.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:08.45 sec
 Dead Time: 0.1%

Verified by:

R. J. J. 12-28-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 22-DEC-1998 16:47:42.41
 Decayed to: 22-DEC-1998 16:47:42.41
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	59.68	264	1.19	120.08	115	11 60.2		AM-241	0.397
0	351.88*	98	1.52	704.14	701	7 95.8			
0	661.61*	19926	1.60	1323.40	1317	14 1.5		CS-137	7.51
0	1172.95	9034	1.92	2346.05	2336	21 2.4		CO-60	4.75
0	1332.16	8201	2.07	2664.55	2655	19 2.3		CO-60	4.78
0	1460.39*	25	2.17	2921.08	2912	16 188.0		K-40	0.148
0	1559.33	29	1.69	3119.05	3114	10 42.7			

Summary of Nuclide Activity

Page : 2

Sample ID : WL27626-STD

Acquisition date : 22-DEC-1998 16:47:42

Total number of lines in spectrum 7
 Number of unidentified lines 1
 Number of lines tentatively identified by NID 6 85.71%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.20E+09Y~~	~~1.00~~	~~1.477E-01~~	~~1.477E-01~~	~~2.775E-01~~	~~187.96~~	~~RAJ~~
CO-60	5.27Y	1.00	4.768E+00	4.768E+00	0.078E+00	1.65	
CS-137	30.00Y	1.00	7.507E+00	7.507E+00	0.114E+00	1.51	
~~AM-241~~	~~432.20Y*****~~	~~1.00~~	~~3.973E-01~~	~~3.973E-01~~	~~2.391E-01~~	~~60.17~~	~~RAJ~~
Total Activity :			1.282E+01	1.282E+01			

Grand Total Activity : 1.282E+01 1.282E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27626-STD

Acquisition date : 22-DEC-1998 16:47:42

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1015.	477.59	3.5637E-01
NA-24	30.	1368.55	1.7757E-02
AR-41	65.	1293.60	3.5679E-02
SC-46	419.	1120.55	4.9427E-02
CR-51	1010.	320.08	2.6762E-01
MN-54	456.	834.83	3.9970E-02
CO-56	515.	846.76	4.2957E-02
CO-58	450.	810.78	3.8947E-02
FE-59	451.	1099.25	8.9223E-02
ZN-65	467.	1115.55	1.0241E-01
SE-75	1274.	264.66	4.4911E-02
KR-85	748.	514.00	7.7543E+00
SR-85	747.	514.01	3.5064E-02
Y-88	11.	1836.06	1.3714E-02
Y-91	153.	1204.67	1.0775E+01
NB-94	547.	871.09	4.5354E-02
NB-95	424.	765.78	3.5939E-02
ZR-95	405.	756.73	6.3671E-02
ZRNB-95	410.	724.18	1.5418E-01
RU-103	772.	497.08	3.6604E-02
RURH-106	468.	621.93	6.3455E-01
AG-108m	404.	722.94	3.6687E-02
CD-109	1063.	88.03	6.3399E-01
AG-110M	954.	657.76	4.9428E-02
SN-113	994.	391.69	4.8658E-02
TE-123m	1171.	159.00	2.0658E-02
SB-124	471.	602.73	3.1450E-02
SB-125	1021.	427.89	1.1511E-01
TE-125m	1167.	109.27	6.5259E+00
I-129	1302.	39.60	0.0000E+00
I-131	963.	364.48	3.5919E-02
XE-131m	1253.	163.93	9.3071E-01
BA-133	939.	356.02	4.5435E-02
CS-134	483.	604.70	3.1966E-02
CS-136	463.	818.51	3.9831E-02
CS-138	33.	1435.86	7.9641E-02
CE-139	1221.	165.85	2.2610E-02
BA-140	593.	537.31	1.2819E-01
LA-140	22.	1596.21	1.7494E-02
CE-141	1264.	145.44	3.6441E-02
CE-144	1213.	133.51	1.5465E-01
CEPR-144	1213.	133.51	3.0901E-01
EU-152	1237.	121.78	6.2428E-02
EU-154	76.	1274.51	7.0504E-02
EU-155	1038.	86.54	7.1027E-02
HF-181	886.	482.18	4.3204E-02
TA-182	987.	67.75	1.0412E-01
HG-203	1181.	279.20	3.2285E-02
BI-207	501.	569.70	3.0873E-02
TL-208	1171.	277.36	4.1268E-01
PB-210	1381.	46.50	0.0000E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27626-STD

Acquisition date : 22-DEC-1998 16:47:42

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	388.	727.18	4.9394E-01
PB-212	1340.	238.63	5.7257E-02
BI-214	610.	609.31	7.8584E-02
PB-214	1042.	351.92	3.2571E-01
RA-224	1338.	240.99	6.3947E-01
RA-226	1380.	186.10	5.8897E-01
AC-228	608.	911.21	1.8641E-01
TH-228	1060.	84.37	1.9498E+00
TH-229	1081.	88.47	9.2640E-02
U-232	1017.	57.78	5.0666E+01
PA-233	1032.	312.17	6.9304E-02
UTH-233	1303.	245.34	2.2865E+01
PA-234M	492.	1001.03	7.4612E+00
TH-234	1020.	63.29	1.5256E+00
U-235	1369.	185.71	3.5654E-02
NP-237	1040.	86.48	1.8838E-01
U-237	1175.	59.54	2.5502E-01
NP-238	452.	984.45	1.6781E-01
NP-239	1103.	106.12	8.1358E-02
PU-239	1193.	129.30	2.6673E+02
AM-243	1062.	74.67	5.4520E-02

*****
 * 222-S Laboratory Counting Room 22-DEC-1998 21:57:39.80
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27626
 Sample ID: WL27626-BLK
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

[Signature]

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3039.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:02.39 sec
 Dead Time: 0.0%

Verified by:

[Signature] 12-23-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 22-DEC-1998 19:27:02.73
 Decayed to: 22-DEC-1998 19:27:02.73
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	46.96	200	1.62	94.66	91	11	50.7		PB-210	0.000E+00
0	351.82*	77	1.14	704.02	698	12	81.5			
0	583.09*	41	1.90	1166.40	1160	12	96.2			
0	609.19*	74	1.41	1218.59	1213	10	57.5			

Summary of Nuclide Activity

Page : 2

Sample ID : WL27626-BLK

Acquisition date : 22-DEC-1998 19:27:02

Total number of lines in spectrum	4	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	4	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~PB-210~~	~~8.14D~~	~~1.00~~	~~0.000E+00~~	~~0.000E+00~~	~~0.000E+00~~	~~0.00~~	<del></del>
Total Activity :			0.000E+00	0.000E+00			

RHJ
12-23-98

Grand Total Activity : 0.000E+00 0.000E+00

Flags: "X" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27626-BLK

Acquisition date : 22-DEC-1998 19:27:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	102.	477.59	5.1644E+00
NA-24	20.	1368.55	6.4573E-01
K-40	371.	1460.75	2.3831E+01
AR-41	30.	1293.60	1.1044E+00
SC-46	67.	1120.55	9.0494E-01
CR-51	158.	320.08	4.7842E+00
MN-54	51.	834.83	6.2031E-01
CO-56	45.	846.76	5.9342E-01
CO-58	47.	810.78	5.8423E-01
FE-59	42.	1099.25	1.2653E+00
CO-60	37.	1332.50	7.9344E-01
ZN-65	33.	1115.55	1.2864E+00
SE-75	207.	264.66	8.1581E-01
KR-85	155.	514.00	1.5914E+02
SR-85	154.	514.01	7.1867E-01
Y-88	6.	1836.06	4.6044E-01
Y-91	34.	1204.67	2.3380E+02
NB-94	46.	871.09	6.1260E-01
NB-95	60.	765.78	6.2108E-01
ZR-95	59.	756.73	1.1220E+00
ZRNB-95	68.	724.18	2.8797E+00
RU-103	90.	497.08	5.7245E-01
RURH-106	75.	621.93	1.1636E+01
AG-108m	64.	722.94	6.6839E-01
CD-109	239.	88.03	1.3494E+01
AG-110M	95.	657.76	7.1536E-01
SN-113	131.	391.69	8.0289E-01
TE-123m	263.	159.00	4.3891E-01
SB-124	71.	602.73	5.5990E-01
SB-125	118.	427.89	1.7818E+00
TE-125m	228.	109.27	1.2961E+02
I-129	472.	39.60	0.0000E+00
I-131	144.	364.48	6.2995E-01
XE-131m	251.	163.93	1.8681E+01
BA-133	134.	356.02	7.7761E-01
CS-134	75.	604.70	5.7577E-01
CS-136	52.	818.51	6.1789E-01
CS-137	434.	661.66	1.6507E+00
CS-138	18.	1435.86	2.6901E+00
CE-139	243.	165.85	4.5259E-01
BA-140	94.	537.31	2.3243E+00
LA-140	12.	1596.21	5.8528E-01
CE-141	247.	145.44	7.2378E-01
CE-144	255.	133.51	3.1806E+00
CEPR-144	255.	133.51	6.3559E+00
EU-152	240.	121.78	1.2362E+00
EU-154	33.	1274.51	2.1171E+00
EU-155	262.	86.54	1.5993E+00
HF-181	102.	482.18	6.6970E-01
TA-182	190.	67.75	2.0568E+00
HG-203	199.	279.20	5.9758E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27626-BLK

Acquisition date : 22-DEC-1998 19:27:02

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	76.	569.70	5.4899E-01
TL-208	197.	277.36	7.6369E+00
BI-212	64.	727.18	9.2020E+00
PB-212	266.	238.63	1.1439E+00
BI-214	220.	609.31	2.1055E+00
PB-214	283.	351.92	7.5811E+00
RA-224	238.	240.99	1.2115E+01
RA-226	283.	186.10	1.1953E+01
AC-228	90.	911.21	3.2734E+00
TH-228	257.	84.37	4.3003E+01
TH-229	242.	88.47	1.9664E+00
U-232	207.	57.78	1.0280E+03
PA-233	181.	312.17	1.3095E+00
UTH-233	198.	245.34	4.0212E+02
PA-234M	35.	1001.03	9.3709E+01
TH-234	202.	63.29	3.0575E+01
U-235	279.	185.71	7.2090E-01
NP-237	263.	86.48	4.2399E+00
U-237	202.	59.54	4.7663E+00
NP-238	41.	984.45	2.3681E+00
NP-239	234.	106.12	1.6804E+00
PU-239	243.	129.30	5.3976E+03
AM-241	202.	59.54	4.5557E+00
AM-243	290.	74.67	1.2745E+00

*****
 * 222-S Laboratory Counting Room 23-DEC-1998 06:22:40.40 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27626
 Sample ID: S98T3384-SAM
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLHabe

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3043.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:36:55.91 sec
 Dead Time: 4.4%

Verified by:

PK June 12-23-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 23-DEC-1998 03:45:14.95
 Decayed to: 23-DEC-1998 03:45:14.95
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:11:20.05

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
10	185.60	1758	1.02	371.75	370	28	40.9	1.91E+01	U-235	69.2
									RA-226	1.139E+03
10	186.98	6053	2.45	374.51	370	28	22.7			
10	189.21	10508	3.29	378.97	370	28	24.0			
0	661.54*	3038394	1.59	1323.25	1314	18	0.1		CS-137	1.964E+05
0	1321.61	871	2.27	2643.44	2635	17	12.9			
0	1508.38	12	0.94	3017.11	3010		15115.6			
0	1763.65*	10	4.10	3527.91	3521		14161.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3384-SAM

Acquisition date : 23-DEC-1998 03:45:14

Total number of lines in spectrum 7
 Number of unidentified lines 4
 Number of lines tentatively identified by NID 3 42.86%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	1.964E+05	1.964E+05	0.002E+05	0.12	
~~RA-226~~	~~1600.00Y*****~~	~~1.00~~	~~1.139E+03~~	~~1.139E+03~~	~~0.466E+03~~	~~40.90~~	~~RA~~
~~U-235~~	~~7.04E+08Y~~	~~1.00~~	~~6.923E+01~~	~~6.923E+01~~	~~2.831E+01~~	~~40.90~~	~~RA~~
Total Activity :			1.976E+05	1.976E+05			12-23-98

Grand Total Activity : 1.976E+05 1.976E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3384-SAM

Acquisition date : 23-DEC-1998 03:45:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	97613.	477.59	6.0856E+02
NA-24	34.	1368.55	3.0903E+00
K-40	384.	1460.75	8.9430E+01
AR-41	377.	1293.60	1.3777E+01
SC-46	643.	1120.55	1.0042E+01
CR-51	86330.	320.08	4.3142E+02
MN-54	1999.	834.83	1.4102E+01
CO-56	1893.	846.76	1.3910E+01
CO-58	2245.	810.78	1.4654E+01
FE-59	673.	1099.25	1.7942E+01
CO-60	61.	1332.50	3.7290E+00
ZN-65	584.	1115.55	1.8825E+01
SE-75	97433.	264.66	6.9800E+01
KR-85	49967.	514.00	1.1074E+04
SR-85	49962.	514.01	5.0092E+01
Y-88	15.	1836.06	2.4550E+00
Y-91	311.	1204.67	2.4806E+03
NB-94	1619.	871.09	1.3189E+01
NB-95	2604.	765.78	1.4995E+01
ZR-95	2690.	756.73	2.7611E+01
ZRNB-95	3249.	724.18	7.3020E+01
RU-103	61190.	497.08	5.6862E+01
RURH-106	18922.	621.93	6.8168E+02
AG-108m	3164.	722.94	1.7269E+01
CD-109	78262.	88.03	1.0154E+03
AG-110M	104930.	657.76	8.7461E+01
SN-113	88722.	391.69	7.8078E+01
TE-123m	97083.	159.00	3.5061E+01
SB-124	19894.	602.73	3.4717E+01
SB-125	101993.	427.89	1.9666E+02
TE-125m	85182.	109.27	9.9723E+03
I-129	70769.	39.60	0.0000E+00
I-131	84920.	364.48	5.7627E+01
XE-131m	98015.	163.93	1.5488E+03
BA-133	84535.	356.02	7.3832E+01
CS-134	19952.	604.70	3.4895E+01
CS-136	2108.	818.51	1.4318E+01
CS-138	21.	1435.86	1.1073E+01
CE-139	97472.	165.85	3.8116E+01
BA-140	35907.	537.31	1.7340E+02
LA-140	18.	1596.21	2.6079E+00
CE-141	97391.	145.44	5.7979E+01
CE-144	97271.	133.51	2.4578E+02
CEPR-144	97272.	133.51	4.9111E+02
EU-152	92950.	121.78	9.5446E+01
EU-154	316.	1274.51	2.2852E+01
EU-155	78195.	86.54	1.1529E+02
HF-181	79223.	482.18	7.1144E+01
TA-182	70755.	67.75	1.6009E+02
HG-203	94367.	279.20	5.1119E+01
BI-207	22671.	569.70	3.5687E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3384-SAM

Acquisition date : 23-DEC-1998 03:45:14

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	94412.	277.36	6.5674E+02
PB-210	69790.	46.50	0.0000E+00
BI-212	3166.	727.18	2.3703E+02
PB-212	105795.	238.63	9.1306E+01
BI-214	19854.	609.31	7.6218E+01
PB-214	84045.	351.92	5.2337E+02
RA-224	104604.	240.99	1.0137E+03
AC-228	1416.	911.21	4.8074E+01
TH-228	77404.	84.37	3.1218E+03
TH-229	78264.	88.47	1.4707E+02
U-232	68988.	57.78	6.4525E+04
PA-233	87197.	312.17	1.1148E+02
UTH-233	102960.	245.34	3.6326E+04
PA-234M	940.	1001.03	1.7170E+03
TH-234	70682.	63.29	2.2176E+03
NP-237	78198.	86.48	3.0541E+02
U-237	69336.	59.54	3.2036E+02
NP-238	1001.	984.45	4.1650E+01
NP-239	82893.	106.12	1.2697E+02
PU-239	95945.	129.30	4.2255E+05
AM-241	69336.	59.54	3.0613E+02
AM-243	74373.	74.67	8.5079E+01

*****
 * 222-S Laboratory Counting Room 23-DEC-1998 09:07:21.81
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27626
 Sample ID: S98T3384-DUP
 Sample Size: 2.50000E-04 L
 Dilution Factor: 1.10000E+01

Removed by:

SLH/aler

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3044.cnf
 Geometry: 43
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:37:06.75 sec
 Dead Time: 4.5%

Verified by:

RJ Jones 12-23-98

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 23-DEC-1998 06:29:47.50
 Decayed to: 23-DEC-1998 06:29:47.50
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:11:20.05

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	74.86	847	0.91	150.41	148	5	84.3		AM-243	56.7
0	84.67	1137	1.48	170.03	168	6	73.4		TH-228	2.738E+03
0	188.47	8779	4.81	377.50	372	15	25.1			
0	661.52*	3115887	1.59	1323.22	1314	18	0.1		CS-137	2.014E+05
0	1321.75	918	2.49	2643.73	2634	19	14.2			
0	1764.11*	12	1.21	3528.83	3522	13143.7				

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3384-DUP

Acquisition date : 23-DEC-1998 06:29:47

Total number of lines in spectrum	6	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	4	66.67%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
CS-137	30.00Y	1.00	2.014E+05	2.014E+05	0.002E+05	0.12			
~~TH-228~~	~~1.91Y~~	~~1.00~~	~~2.737E+03~~	~~2.738E+03~~	~~2.009E+03~~	~~73.40~~			<i>raj</i>
~~AM-243~~	~~7380.00Y*****~~		~~5.673E+01~~	~~5.673E+01~~	~~4.780E+01~~	~~84.27~~			<i>raj</i>
Total Activity :			2.042E+05	2.042E+05					12-23-88

Grand Total Activity :	2.042E+05	2.042E+05

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3384-DUP

Acquisition date : 23-DEC-1998 06:29:47

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	100382.	477.59	6.1712E+02
NA-24	37.	1368.55	3.2027E+00
K-40	369.	1460.75	8.7660E+01
AR-41	385.	1293.60	1.3935E+01
SC-46	675.	1120.55	1.0281E+01
CR-51	88156.	320.08	4.3595E+02
MN-54	2102.	834.83	1.4460E+01
CO-56	2102.	846.76	1.4645E+01
CO-58	2251.	810.78	1.4674E+01
FE-59	740.	1099.25	1.8792E+01
CO-60	68.	1332.50	3.9272E+00
ZN-65	662.	1115.55	2.0022E+01
SE-75	99979.	264.66	7.0704E+01
KR-85	51638.	514.00	1.1258E+04
SR-85	51629.	514.01	5.0919E+01
Y-88	13.	1836.06	2.3039E+00
Y-91	316.	1204.67	2.4986E+03
NB-94	1823.	871.09	1.3986E+01
NB-95	2640.	765.78	1.5094E+01
ZR-95	2845.	756.73	2.8385E+01
ZRNB-95	3394.	724.18	7.4614E+01
RU-103	62797.	497.08	5.7603E+01
RURH-106	19383.	621.93	6.8989E+02
AG-108m	3398.	722.94	1.7888E+01
CD-109	79470.	88.03	1.0232E+03
AG-110M	111462.	657.76	9.0138E+01
SN-113	91177.	391.69	7.9149E+01
TE-123m	99263.	159.00	3.5452E+01
SB-124	20435.	602.73	3.5184E+01
SB-125	104973.	427.89	1.9951E+02
TE-125m	87294.	109.27	1.0095E+04
I-129	72688.	39.60	0.0000E+00
I-131	86876.	364.48	5.8286E+01
XE-131m	100241.	163.93	1.5663E+03
BA-133	86791.	356.02	7.4808E+01
CS-134	20312.	604.70	3.5207E+01
CS-136	2170.	818.51	1.4525E+01
CS-138	31.	1435.86	1.3101E+01
CE-139	100173.	165.85	3.8639E+01
BA-140	36940.	537.31	1.7587E+02
LA-140	25.	1596.21	2.9887E+00
CE-141	100274.	145.44	5.8829E+01
CE-144	99638.	133.51	2.4874E+02
CEPR-144	99640.	133.51	4.9704E+02
EU-152	95610.	121.78	9.6800E+01
EU-154	327.	1274.51	2.3245E+01
EU-155	79415.	86.54	1.1619E+02
HF-181	81097.	482.18	7.1978E+01
TA-182	73323.	67.75	1.6297E+02
HG-203	96130.	279.20	5.1594E+01
BI-207	23329.	569.70	3.6199E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3384-DUP

Acquisition date : 23-DEC-1998 06:29:47

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	96981.	277.36	6.6560E+02
PB-210	71283.	46.50	0.0000E+00
BI-212	3309.	727.18	2.4230E+02
PB-212	107707.	238.63	9.2126E+01
BI-214	19992.	609.31	7.6482E+01
PB-214	86076.	351.92	5.3021E+02
RA-224	106745.	240.99	1.0240E+03
RA-226	130024.	186.10	1.0878E+03
AC-228	1417.	911.21	4.8091E+01
TH-229	79573.	88.47	1.4829E+02
U-232	71690.	57.78	6.5774E+04
PA-233	89288.	312.17	1.1280E+02
UTH-233	105539.	245.34	3.6777E+04
PA-234M	999.	1001.03	1.7691E+03
TH-234	71415.	63.29	2.2290E+03
U-235	128158.	185.71	6.5652E+01
NP-237	79430.	86.48	3.0780E+02
U-237	71722.	59.54	3.2582E+02
NP-238	1063.	984.45	4.2903E+01
NP-239	85374.	106.12	1.2886E+02
PU-239	98729.	129.30	4.2862E+05
AM-241	71722.	59.54	3.1134E+02

LABCORE Completed Worklist Report for Worklist# 27865

Analyst: gll

Instrument: GEA03

Book#: 46B57

Method: LA-548-121 Rev/Mod _____

Worklist Comment: BY105 FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@GEA-01	C060-02	SOLID	5.508e-03	6.05e-03	109.840	% Recovery
1 STD	0	@GEA-01	C060-02E	SOLID	1	2.51	2.510	% Ct Error
1 STD	0	@GEA-01	CS13702	SOLID	8.151e-03	7.94e-03	97.411	% Recovery
1 STD	0	@GEA-01	CS13702E	SOLID	1	2.63	2.630	% Ct Error
2 BLNK-PREP	0	@GEA-01	C060-02	SOLID	1	<7.49e-3		uCi/g
2 BLNK-PREP	0	@GEA-01	CS13702	SOLID	1	<1.09e-2		uCi/g
3 SAMPLE	S98T003393	0 F	@GEA-01	C060-02	SOLID	N/A	6.975e-03	uCi/g
3 SAMPLE	S98T003393	0 F	@GEA-01	C060-02E	SOLID	N/A		% Ct. Error
3 SAMPLE	S98T003393	0 F	@GEA-01	CS13702	SOLID	N/A	8.342e+00	uCi/g
3 SAMPLE	S98T003393	0 F	@GEA-01	CS13702E	SOLID	N/A	0.970	% Ct. Error
4 DUP	S98T003393	0 F	@GEA-01	C060-02	SOLID	<6.98e-3	<7.86e-3	RPD
4 DUP	S98T003393	0 F	@GEA-01	C060-02E	SOLID	1	n/a	% Ct Error
4 DUP	S98T003393	0 F	@GEA-01	CS13702	SOLID	8.34e+00	8.86e+00	6.047 RPD
4 DUP	S98T003393	0 F	@GEA-01	CS13702E	SOLID	1.00	0.940	0.940 % Ct Error
5 SAMPLE	S98T003398	0 F	@GEA-01	C060-02	SOLID	N/A	<7.998e-01	0.800 uCi/g
5 SAMPLE	S98T003398	0 F	@GEA-01	C060-02E	SOLID	N/A	n/a	% Ct. Error
5 SAMPLE	S98T003398	0 F	@GEA-01	CS13702	SOLID	N/A	2.331e+02	uCi/g
5 SAMPLE	S98T003398	0 F	@GEA-01	CS13702E	SOLID	N/A	1.91	% Ct. Error
6 DUP	S98T003398	0 F	@GEA-01	C060-02	SOLID	<8.00e-1	<6.82e-1	RPD
6 DUP	S98T003398	0 F	@GEA-01	C060-02E	SOLID	1	n/a	% Ct Error
6 DUP	S98T003398	0 F	@GEA-01	CS13702	SOLID	2.33e+02	2.30e+02	1.296 RPD
6 DUP	S98T003398	0 F	@GEA-01	CS13702E	SOLID	1.00	1.92	1.920 % Ct Error
7 SAMPLE	S98T003388	0 F	@GEA-01	C060-02	SOLID	N/A	8.815e-03	8.81e-003 uCi/g
7 SAMPLE	S98T003388	0 F	@GEA-01	C060-02E	SOLID	N/A	n/a	% Ct. Error
7 SAMPLE	S98T003388	0 F	@GEA-01	CS13702	SOLID	N/A	1.652e+01	uCi/g
7 SAMPLE	S98T003388	0 F	@GEA-01	CS13702E	SOLID	N/A	0.690	% Ct. Error
8 DUP	S98T003388	0 F	@GEA-01	C060-02	SOLID	<8.81e-3	<7.26e-3	RPD
8 DUP	S98T003388	0 F	@GEA-01	C060-02E	SOLID	1	n/a	% Ct Error
8 DUP	S98T003388	0 F	@GEA-01	CS13702	SOLID	1.65e+01	1.58e+01	4.334 RPD
8 DUP	S98T003388	0 F	@GEA-01	CS13702E	SOLID	1.00	0.700	0.700 % Ct Error

Comments Section:

Comments for sample# S98T003393 and test @GEA-01 .

DL=0 => n/a.

Comments for sample# S98T003398 and test @GEA-01 .

DL=0 => n/a.

Comments for sample# S98T003388 and test @GEA-01 .

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 27865

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
		DL=0	=	>						n/a.

Final page for worklist# 27865

_____
Analyst Signature   Date

_____
Analyst Signature   Date

John Pelyea 15 Jan 99
_____
Reviewer Signature   Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27865

Analyst: gll Instrument: GEA00 +3 Book# 46857Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 FOR @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	SOLID		
2	BLNK-PREP				@GEA-01	SOLID		
3	SAMPLE	S98T003393	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
4	DUP	S98T003393	0	F	@GEA-01	SOLID		
5	SAMPLE	S98T003398	0	F	@GEA-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
6	DUP	S98T003398	0	F	@GEA-01	SOLID		
7	SAMPLE	S98T003388	0	F	@GEA-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E						
8	DUP	S98T003388	0	F	@GEA-01	SOLID		

Final page for worklist # 27865

Analyst Signature *gll* Date 1-14-99Analyst Signature *M. Brown* Date 1/15/99

Data Entry Comments:

S = Worklist Slot Number, *R* = Replicate Number, *A* = Aliquot Code.

*****
 * 222-S Laboratory Counting Room 14-JAN-1999 10:43:59.74 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27865
 Sample ID: WL27865-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3134.cnf
 Geometry: 42
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:03.13 sec
 Dead Time: 0.1%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-JAN-1999 09:53:25.68
 Decayed to: 14-JAN-1999 09:53:25.68
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.76*	7025	1.56	1323.69	1317	15	2.6		CS-137	7.94
0	1173.15	3786	1.97	2346.44	2338	19	3.7		CO-60	5.97
0	1332.34	3496	2.09	2664.91	2654	22	3.5		CO-60	6.12
0	1460.67*	15	2.19	2921.65	2914	14171.3			K-40	0.263

Summary of Nuclide Activity
Sample ID : WL27865-STD

Acquisition date : 14-JAN-1999 09:53:25

Total number of lines in spectrum	4	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	4	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~2.627E-01~~	~~2.627E-01~~	~~4.501E-01~~	~~171.32~~	<del></del>
CO-60	5.27Y	1.00	6.047E+00	6.047E+00	0.152E+00	2.51	
CS-137	30.00Y	1.00	7.940E+00	7.940E+00	0.209E+00	2.63	
Total Activity :			1.425E+01	1.425E+01			

JJO
H44

Grand Total Activity : 1.425E+01 1.425E+01

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	398.	477.59	6.7606E-01
NA-24	12.	1368.55	3.4256E-02
AR-41	24.	1293.60	5.1359E-02
SC-46	174.	1120.55	9.7014E-02
CR-51	398.	320.08	5.0873E-01
MN-54	212.	834.83	8.2840E-02
CO-56	214.	846.76	8.4270E-02
CO-58	193.	810.78	7.7675E-02
FE-59	180.	1099.25	1.7157E-01
ZN-65	181.	1115.55	1.9420E-01
SE-75	463.	264.66	8.2048E-02
KR-85	269.	514.00	1.4141E+01
SR-85	268.	514.01	6.3895E-02
Y-88	7.	1836.06	3.3476E-02
Y-91	62.	1204.67	2.1118E+01
NB-94	225.	871.09	8.8329E-02
NB-95	170.	765.78	6.9291E-02
ZR-95	180.	756.73	1.2906E-01
ZRNB-95	168.	724.18	3.0027E-01
RU-103	302.	497.08	6.9472E-02
RURH-106	193.	621.93	1.2415E+00
AG-108m	176.	722.94	7.3750E-02
CD-109	422.	88.03	1.2103E+00
AG-110M	304.	657.76	8.4839E-02
SN-113	349.	391.69	8.7527E-02
TE-123m	474.	159.00	3.9787E-02
SB-124	193.	602.73	6.1217E-02
SB-125	380.	427.89	2.1308E-01
TE-125m	454.	109.27	1.2336E+01
I-129	459.	39.60	0.0000E+00
I-131	323.	364.48	6.3046E-02
XE-131m	480.	163.93	1.7409E+00
BA-133	365.	356.02	8.5904E-02
CS-134	200.	604.70	6.2574E-02
CS-136	196.	818.51	7.8645E-02
CS-138	9.	1435.86	6.6291E-02
CE-139	467.	165.85	4.2375E-02
BA-140	219.	537.31	2.3687E-01
LA-140	5.	1596.21	2.8723E-02
CE-141	472.	145.44	6.7420E-02
CE-144	461.	133.51	2.8876E-01
CEPR-144	461.	133.51	5.7701E-01
EU-152	431.	121.78	1.1180E-01
EU-154	30.	1274.51	1.3719E-01
EU-155	421.	86.54	1.3703E-01
HF-181	336.	482.18	8.0693E-02
TA-182	386.	67.75	1.9752E-01
HG-203	451.	279.20	6.0405E-02
BI-207	197.	569.70	5.8934E-02
TL-208	436.	277.36	7.6330E-01
PB-210	487.	46.50	0.0000E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27865-STD

Acquisition date : 14-JAN-1999 09:53:25

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-212	153.	727.18	9.4510E-01
PB-212	554.	238.63	1.1140E-01
BI-214	212.	609.31	1.4108E-01
PB-214	384.	351.92	2.6986E-01
RA-224	535.	240.99	1.2161E+00
RA-226	532.	186.10	1.1074E+00
AC-228	259.	911.21	3.6913E-01
TH-228	392.	84.37	3.5972E+00
TH-229	420.	88.47	1.7502E-01
U-232	389.	57.78	9.5098E+01
PA-233	405.	312.17	1.3158E-01
UTH-233	518.	245.34	4.3655E+01
PA-234M	209.	1001.03	1.4767E+01
TH-234	372.	63.29	2.7984E+00
U-235	525.	185.71	6.6886E-02
NP-237	419.	86.48	3.6224E-01
U-237	370.	59.54	4.3313E-01
NP-238	221.	984.45	3.5225E-01
NP-239	407.	106.12	1.4832E-01
PU-239	463.	129.30	5.0365E+02
AM-241	370.	59.54	4.1544E-01
AM-243	456.	74.67	1.0819E-01

*****

* 222-S Laboratory Counting Room 14-JAN-1999 13:29:46.44 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27865
 Sample ID: WL27865-BLNK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

NE Wright

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3135.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:02.26 sec
 Dead Time: 0.0%

Verified by:

John Relyea 15 Jan 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 14-JAN-1999 10:59:10.87
 Decayed to: 14-JAN-1999 10:59:10.87
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: MB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	661.80*	58	1.54	1323.77	1316	15100.0		CS-137	2.203E-02

Summary of Nuclide Activity

Page : 2

Sample ID : WL27865-BLNK

Acquisition date : 14-JAN-1999 10:59:10

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~2.203E-02~~	~~2.203E-02~~	~~2.203E-02~~	~~100.00~~	~~1-54~~
Total Activity :			2.203E-02	2.203E-02			

Grand Total Activity : 2.203E-02   2.203E-02

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL27865-BLNK

Acquisition date : 14-JAN-1999 10:59:10

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	98.	477.59	1.1495E-01
NA-24	19.	1368.55	1.4444E-02
K-40	372.	1460.75	5.4195E-01
AR-41	22.	1293.60	2.1939E-02
SC-46	71.	1120.55	2.1176E-02
CR-51	159.	320.08	1.0918E-01
MN-54	43.	834.83	1.3014E-02
CO-56	58.	846.76	1.5139E-02
CO-58	56.	810.78	1.4446E-02
FE-59	29.	1099.25	2.4429E-02
CO-60	25.	1332.50	1.5085E-02
ZN-65	28.	1115.55	2.7093E-02
SE-75	220.	264.66	1.9073E-02
KR-85	172.	514.00	3.8081E+00
SR-85	172.	514.01	1.7196E-02
Y-88	7.	1836.06	1.1261E-02
Y-91	49.	1204.67	6.2916E+00
NB-94	37.	871.09	1.2652E-02
NB-95	59.	765.78	1.4008E-02
ZR-95	54.	756.73	2.4335E-02
ZRNB-95	68.	724.18	6.5467E-02
RU-103	80.	497.08	1.2264E-02
RURH-106	87.	621.93	2.8255E-01
AG-108m	64.	722.94	1.5267E-02
CD-109	221.	88.03	2.9483E-01
AG-110M	69.	657.76	1.3950E-02
SN-113	142.	391.69	1.8952E-02
TE-123m	230.	159.00	9.3562E-03
SB-124	85.	602.73	1.3835E-02
SB-125	126.	427.89	4.1818E-02
TE-125m	243.	109.27	3.0347E+00
I-129	461.	39.60	0.0000E+00
I-131	130.	364.48	1.3601E-02
XE-131m	245.	163.93	4.1981E-01
BA-133	138.	356.02	1.7914E-02
CS-134	77.	604.70	1.3246E-02
CS-136	44.	818.51	1.3002E-02
CS-138	19.	1435.86	6.2608E-02
CE-139	240.	165.85	1.0226E-02
BA-140	85.	537.31	5.0499E-02
LA-140	16.	1596.21	1.5116E-02
CE-141	243.	145.44	1.6307E-02
CE-144	242.	133.51	7.0417E-02
CEPR-144	242.	133.51	1.4074E-01
EU-152	231.	121.78	2.7550E-02
EU-154	30.	1274.51	4.6160E-02
EU-155	230.	86.54	3.4112E-02
HF-181	97.	482.18	1.4867E-02
TA-182	199.	67.75	4.7748E-02
HG-203	181.	279.20	1.2953E-02
BI-207	82.	569.70	1.2984E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL27865-BLNK

Acquisition date : 14-JAN-1999 10:59:10

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	184.	277.36	1.6780E-01
PB-210	411.	46.50	0.0000E+00
BI-212	72.	727.18	2.2117E-01
PB-212	299.	238.63	2.7514E-02
BI-214	127.	609.31	3.6768E-02
PB-214	232.	351.92	1.5672E-01
RA-224	260.	240.99	2.8746E-01
RA-226	295.	186.10	2.7698E-01
AC-228	90.	911.21	7.4180E-02
TH-228	248.	84.37	9.6068E-01
TH-229	220.	88.47	4.2700E-02
U-232	180.	57.78	2.1840E+01
PA-233	181.	312.17	2.9704E-02
UTH-233	231.	245.34	9.8341E+00
PA-234M	42.	1001.03	2.3211E+00
TH-234	206.	63.29	7.0037E-01
U-235	288.	185.71	1.6644E-02
NP-237	231.	86.48	9.0497E-02
U-237	198.	59.54	1.0720E-01
NP-238	53.	984.45	6.0648E-02
NP-239	238.	106.12	3.8542E-02
PU-239	241.	129.30	1.2240E+02
AM-241	198.	59.54	1.0246E-01
AM-243	273.	74.67	2.8092E-02

*****
 * 222-S Laboratory Counting Room 14-JAN-1999 17:41:37.58 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3393-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Sheld

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3137.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:07.70 sec
 Dead Time: 0.1%

Verified by:

John Relyea 15 Jan 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 14-JAN-1999 15:11:01.01
 Decayed to: 14-JAN-1999 15:11:01.01
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SZC
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	661.80*	44591	1.60	1323.78	1315	18 1.0		CS-137	16.8
0	1460.71*	30	2.22	2921.72	2911	18153.4		K-40	0.177

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3393-SAM

Acquisition date : 14-JAN-1999 15:11:01

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.766E-01~~	~~1.766E-01~~	~~2.709E-01~~	~~153.41~~	~~820~~
CS-137	30.00Y	1.00	1.680E+01	1.680E+01	0.016E+01	0.97	154
Total Activity :			1.698E+01	1.698E+01			

Grand Total Activity :	1.698E+01	1.698E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3393-SAM

Acquisition date : 14-JAN-1999 15:11:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1595.	477.59	4.4510E-01
NA-24	25.	1368.55	1.6323E-02
AR-41	22.	1293.60	2.1884E-02
SC-46	57.	1120.55	1.9109E-02
CR-51	1319.	320.08	3.0520E-01
MN-54	57.	834.83	1.4789E-02
CO-56	55.	846.76	1.4746E-02
CO-58	40.	810.78	1.2344E-02
FE-59	46.	1099.25	3.0056E-02
CO-60	21.	1332.50	1.4048E-02
ZN-65	31.	1115.55	2.8463E-02
SE-75	1431.	264.66	4.7559E-02
KR-85	873.	514.00	8.3637E+00
SR-85	872.	514.01	3.7817E-02
Y-88	2.	1836.06	6.9804E-03
Y-91	42.	1204.67	5.8869E+00
NB-94	57.	871.09	1.5320E-02
NB-95	58.	765.78	1.3945E-02
ZR-95	59.	756.73	2.5357E-02
ZRNB-95	73.	724.18	6.7649E-02
RU-103	976.	497.08	4.1064E-02
RURH-106	319.	621.93	5.2714E-01
AG-108m	73.	722.94	1.6147E-02
CD-109	1317.	88.03	7.0424E-01
AG-110M	925.	657.76	4.8680E-02
SN-113	1358.	391.69	5.6743E-02
TE-123m	1548.	159.00	2.3696E-02
SB-124	329.	602.73	2.6417E-02
SB-125	1596.	427.89	1.4341E-01
TE-125m	1413.	109.27	7.1689E+00
I-129	1440.	39.60	0.0000E+00
I-131	1293.	364.48	4.1527E-02
XE-131m	1513.	163.93	1.0210E+00
BA-133	1301.	356.02	5.3331E-02
CS-134	342.	604.70	2.7048E-02
CS-136	45.	818.51	1.3165E-02
CS-138	24.	1435.86	6.8871E-02
CE-139	1483.	165.85	2.4880E-02
BA-140	555.	537.31	1.2407E-01
LA-140	8.	1596.21	1.1548E-02
CE-141	1445.	145.44	3.8924E-02
CE-144	1505.	133.51	1.7196E-01
CEPR-144	1505.	133.51	3.4355E-01
EU-152	1476.	121.78	6.8118E-02
EU-154	32.	1274.51	4.7686E-02
EU-155	1331.	86.54	8.0290E-02
HF-181	1224.	482.18	5.0641E-02
TA-182	1137.	67.75	1.1160E-01
HG-203	1457.	279.20	3.5794E-02
BI-207	359.	569.70	2.6251E-02
TL-208	1428.	277.36	4.5509E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3393-SAM

Acquisition date : 14-JAN-1999 15:11:01

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	1444.	46.50	0.0000E+00
BI-212	60.	727.18	2.0300E-01
PB-212	1735.	238.63	6.5035E-02
BI-214	377.	609.31	6.2114E-02
PB-214	1400.	351.92	3.7653E-01
RA-224	1686.	240.99	7.1659E-01
RA-226	1926.	186.10	6.9400E-01
AC-228	72.	911.21	6.7006E-02
TH-228	1309.	84.37	2.1623E+00
TH-229	1321.	88.47	1.0225E-01
U-232	1136.	57.78	5.3498E+01
PA-233	1317.	312.17	7.8130E-02
UTH-233	1556.	245.34	2.4953E+01
PA-234M	36.	1001.03	2.1669E+00
TH-234	1169.	63.29	1.6316E+00
U-235	1901.	185.71	4.1906E-02
NP-237	1331.	86.48	2.1262E-01
U-237	1171.	59.54	2.5462E-01
NP-238	39.	984.45	5.2665E-02
NP-239	1296.	106.12	8.8074E-02
PU-239	1456.	129.30	2.9416E+02
AM-241	1171.	59.54	2.4335E-01
AM-243	1331.	74.67	6.0910E-02

*****
 * 222-S Laboratory Counting Room 14-JAN-1999 20:23:21.42 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3393-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

gldale

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3138.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:08.11 sec
 Dead Time: 0.1%

Verified by:

John Belger 15 Jan 99

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 14-JAN-1999 17:52:40.71
 Decayed to: 14-JAN-1999 17:52:40.71
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.80*	47720	1.61	1323.78	1315	18	0.9		CS-137	18.0
0	1120.83*	23	1.98	2241.79	2235	121	16.5			
0	1764.32*	21	1.63	3529.26	3522	14	87.2			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3393-DUP

Acquisition date : 14-JAN-1999 17:52:40

Total number of lines in spectrum	3	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	3	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean	Wtd Mean	Decay Corr	2-Sigma	2-Sigma Error	%Error	Flags
			Uncorrected	Decay Corr					
			uCi/L	uCi/L					
CS-137	30.00Y	1.00	1.798E+01	1.798E+01	0.017E+01	0.94			
Total Activity :			1.798E+01	1.798E+01					

Grand Total Activity : 1.798E+01 1.798E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3393-DUP

Acquisition date : 14-JAN-1999 17:52:40

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	1652.	477.59	4.5299E-01
NA-24	20.	1368.55	1.4778E-02
K-40	371.	1460.75	5.4123E-01
AR-41	26.	1293.60	2.3596E-02
SC-46	72.	1120.55	2.1297E-02
CR-51	1438.	320.08	3.1841E-01
MN-54	46.	834.83	1.3407E-02
CO-56	69.	846.76	1.6385E-02
CO-58	38.	810.78	1.2052E-02
FE-59	37.	1099.25	2.7189E-02
CO-60	28.	1332.50	1.5966E-02
ZN-65	41.	1115.55	3.2097E-02
SE-75	1562.	264.66	4.9659E-02
KR-85	875.	514.00	8.3762E+00
SR-85	875.	514.01	3.7879E-02
Y-88	12.	1836.06	1.4227E-02
Y-91	43.	1204.67	5.9679E+00
NB-94	48.	871.09	1.4232E-02
NB-95	65.	765.78	1.4707E-02
ZR-95	57.	756.73	2.5019E-02
ZRNB-95	58.	724.18	6.0770E-02
RU-103	1064.	497.08	4.2832E-02
RURH-106	343.	621.93	5.4564E-01
AG-108m	57.	722.94	1.4391E-02
CD-109	1384.	88.03	7.2176E-01
AG-110M	984.	657.76	5.0177E-02
SN-113	1462.	391.69	5.8837E-02
TE-123m	1570.	159.00	2.3865E-02
SB-124	343.	602.73	2.6954E-02
SB-125	1703.	427.89	1.4805E-01
TE-125m	1379.	109.27	7.0831E+00
I-129	1499.	39.60	0.0000E+00
I-131	1383.	364.48	4.2918E-02
XE-131m	1569.	163.93	1.0397E+00
BA-133	1301.	356.02	5.3327E-02
CS-134	369.	604.70	2.8067E-02
CS-136	49.	818.51	1.3689E-02
CS-138	23.	1435.86	6.7491E-02
CE-139	1612.	165.85	2.5926E-02
BA-140	636.	537.31	1.3263E-01
LA-140	15.	1596.21	1.4862E-02
CE-141	1615.	145.44	4.1110E-02
CE-144	1600.	133.51	1.7722E-01
CEPR-144	1600.	133.51	3.5415E-01
EU-152	1545.	121.78	6.9656E-02
EU-154	37.	1274.51	5.0594E-02
EU-155	1427.	86.54	8.3075E-02
HF-181	1266.	482.18	5.1480E-02
TA-182	1238.	67.75	1.1638E-01
HG-203	1394.	279.20	3.5023E-02
BI-207	397.	569.70	2.7546E-02

4627

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3393-DUP

Acquisition date : 14-JAN-1999 17:52:40

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	1409.	277.36	4.5210E-01
PB-210	1504.	46.50	0.0000E+00
BI-212	63.	727.18	2.0760E-01
PB-212	1820.	238.63	6.6581E-02
BI-214	442.	609.31	6.7171E-02
PB-214	1387.	351.92	3.7489E-01
RA-224	1700.	240.99	7.1963E-01
RA-226	1976.	186.10	7.0280E-01
AC-228	82.	911.21	7.1158E-02
TH-228	1375.	84.37	2.2160E+00
TH-229	1336.	88.47	1.0280E-01
U-232	1211.	57.78	5.5215E+01
PA-233	1332.	312.17	7.8571E-02
UTH-233	1699.	245.34	2.6053E+01
PA-234M	51.	1001.03	2.5413E+00
TH-234	1287.	63.29	1.7104E+00
U-235	1939.	185.71	4.2322E-02
NP-237	1427.	86.48	2.2008E-01
U-237	1272.	59.54	2.6520E-01
NP-238	43.	984.45	5.5005E-02
NP-239	1337.	106.12	8.9422E-02
PU-239	1511.	129.30	2.9956E+02
AM-241	1272.	59.54	2.5347E-01
AM-243	1283.	74.67	5.9830E-02

*****
 * 222-S Laboratory Counting Room 14-JAN-1999 23:05:14.70 *
 *****

>>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3398-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

gush

>>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3139.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:03.81 sec
 Dead Time: 0.0%

Verified by:

John Reizer 15 Jan 99

>>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 14-JAN-1999 20:34:40.93
 Decayed to: 14-JAN-1999 20:34:40.93
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uci/L
0	661.81*	12168	1.59	1323.80	1315	17	1.9		CS-137	463.
0	911.33*	45	2.36	1822.78	1817	12	63.7			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3398-SAM

Acquisition date : 14-JAN-1999 20:34:40

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	4.630E+02	4.630E+02	0.089E+02	1.91	
Total Activity :			4.630E+02	4.630E+02			

Grand Total Activity :	4.630E+02	4.630E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3398-SAM

Acquisition date : 14-JAN-1999 20:34:40

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	456.	477.59	2.4328E+01
NA-24	23.	1368.55	1.5886E+00
K-40	363.	1460.75	5.4116E+01
AR-41	32.	1293.60	2.5905E+00
SC-46	68.	1120.55	2.0912E+00
CR-51	486.	320.08	1.8911E+01
MN-54	49.	834.83	1.3947E+00
CO-56	46.	846.76	1.3709E+00
CO-58	44.	810.78	1.3002E+00
FE-59	39.	1099.25	2.8345E+00
CO-60	27.	1332.50	1.5885E+00
ZN-65	43.	1115.55	3.3383E+00
SE-75	564.	264.66	3.0439E+00
KR-85	351.	514.00	5.4164E+02
SR-85	350.	514.01	2.4484E+00
Y-88	12.	1836.06	1.4174E+00
Y-91	38.	1204.67	5.6982E+02
NB-94	45.	871.09	1.3918E+00
NB-95	47.	765.78	1.2771E+00
ZR-95	53.	756.73	2.4472E+00
ZRNB-95	50.	724.18	5.7375E+00
RU-103	337.	497.08	2.4675E+00
RURH-106	156.	621.93	3.7669E+01
AG-108m	48.	722.94	1.3393E+00
CD-109	665.	88.03	5.0863E+01
AG-110M	313.	657.76	2.8992E+00
SN-113	442.	391.69	3.3077E+00
TE-123m	630.	159.00	1.5393E+00
SB-124	148.	602.73	1.8146E+00
SB-125	526.	427.89	8.4005E+00
TE-125m	589.	109.27	4.7136E+02
I-129	784.	39.60	0.0000E+00
I-131	438.	364.48	2.4679E+00
XE-131m	647.	163.93	6.7953E+01
BA-133	466.	356.02	3.2579E+00
CS-134	130.	604.70	1.7165E+00
CS-136	63.	818.51	1.5489E+00
CS-138	23.	1435.86	6.8889E+00
CE-139	647.	165.85	1.6723E+00
BA-140	215.	537.31	7.9225E+00
LA-140	15.	1596.21	1.5210E+00
CE-141	648.	145.44	2.6524E+00
CE-144	660.	133.51	1.1589E+01
CEPR-144	660.	133.51	2.3158E+01
EU-152	606.	121.78	4.4439E+00
EU-154	30.	1274.51	4.6341E+00
EU-155	635.	86.54	5.6411E+00
HF-181	393.	482.18	2.9351E+00
TA-182	538.	67.75	7.8138E+00
HG-203	483.	279.20	2.1043E+00
BI-207	166.	569.70	1.8299E+00

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3398-SAM

Acquisition date : 14-JAN-1999 20:34:40

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	489.	277.36	2.7178E+01
PB-210	698.	46.50	0.0000E+00
BI-212	70.	727.18	2.2039E+01
PB-212	696.	238.63	4.1939E+00
BI-214	213.	609.31	4.7603E+00
PB-214	539.	351.92	2.3822E+01
RA-224	642.	240.99	4.5050E+01
RA-226	724.	186.10	4.3333E+01
AC-228	100.	911.21	7.8756E+00
TH-228	621.	84.37	1.5148E+02
TH-229	670.	88.47	7.3981E+00
U-232	530.	57.78	3.7189E+03
PA-233	458.	312.17	4.7029E+00
UTH-233	610.	245.34	1.5919E+03
PA-234M	35.	1001.03	2.1433E+02
TH-234	512.	63.29	1.1001E+02
U-235	730.	185.71	2.6449E+00
NP-237	633.	86.48	1.4914E+01
U-237	547.	59.54	1.7717E+01
NP-238	45.	984.45	5.6705E+00
NP-239	610.	106.12	6.1460E+00
PU-239	660.	129.30	2.0144E+04
AM-241	547.	59.54	1.6933E+01
AM-243	638.	74.67	4.2887E+00

*****
 * 222-S Laboratory Counting Room 15-JAN-1999 04:21:53.40 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3398-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.01000E+02

Removed by:

John Relyea

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3142.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:03.78 sec
 Dead Time: 0.0%

Verified by:

John Relyea 15 Jan 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 15-JAN-1999 01:51:19.77
 Decayed to: 15-JAN-1999 01:51:19.77
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	661.82*	12020	1.56	1323.81	1317	15 1.9		CS-137	457.
0	1460.73*	30	2.07	2921.78	2914	15146.9		K-40	18.1

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3398-DUP

Acquisition date : 15-JAN-1999 01:51:19

Total number of lines in spectrum	2	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	2	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~1.807E+01~~	~~1.807E+01~~	~~2.654E+01~~	~~146.91~~	<del></del>
CS-137	30.00Y	1.00	4.574E+02	4.574E+02	0.088E+02	1.92	
Total Activity :			4.754E+02	4.754E+02			

Grand Total Activity : 4.754E+02 4.754E+02

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3398-DUP

Acquisition date : 15-JAN-1999 01:51:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	472.	477.59	2.4745E+01
NA-24	27.	1368.55	1.7156E+00
AR-41	27.	1293.60	2.3974E+00
SC-46	63.	1120.55	2.0230E+00
CR-51	474.	320.08	1.8675E+01
MN-54	52.	834.83	1.4297E+00
CO-56	65.	846.76	1.6115E+00
CO-58	45.	810.78	1.3108E+00
FE-59	42.	1099.25	2.9282E+00
CO-60	19.	1332.50	1.3536E+00
ZN-65	48.	1115.55	3.5033E+00
SE-75	551.	264.66	3.0078E+00
KR-85	351.	514.00	5.4182E+02
SR-85	351.	514.01	2.4489E+00
Y-88	9.	1836.06	1.2629E+00
Y-91	30.	1204.67	5.0624E+02
NB-94	35.	871.09	1.2342E+00
NB-95	61.	765.78	1.4332E+00
ZR-95	57.	756.73	2.5310E+00
ZRNB-95	74.	724.18	6.8745E+00
RU-103	302.	497.08	2.3417E+00
RURH-106	136.	621.93	3.5345E+01
AG-108m	72.	722.94	1.6263E+00
CD-109	663.	88.03	5.0813E+01
AG-110M	290.	657.76	2.7955E+00
SN-113	488.	391.69	3.4699E+00
TE-123m	580.	159.00	1.4785E+00
SB-124	137.	602.73	1.7525E+00
SB-125	476.	427.89	8.0020E+00
TE-125m	631.	109.27	4.8766E+02
I-129	751.	39.60	0.0000E+00
I-131	439.	364.48	2.4711E+00
XE-131m	623.	163.93	6.6742E+01
BA-133	495.	356.02	3.3537E+00
CS-134	143.	604.70	1.7981E+00
CS-136	38.	818.51	1.2229E+00
CS-138	26.	1435.86	7.2818E+00
CE-139	646.	165.85	1.6717E+00
BA-140	229.	537.31	8.1662E+00
LA-140	10.	1596.21	1.2680E+00
CE-141	673.	145.44	2.7022E+00
CE-144	652.	133.51	1.1522E+01
CEPR-144	652.	133.51	2.3024E+01
EU-152	640.	121.78	4.5633E+00
EU-154	39.	1274.51	5.2325E+00
EU-155	668.	86.54	5.7794E+00
HF-181	402.	482.18	2.9656E+00
TA-182	566.	67.75	8.0099E+00
HG-203	532.	279.20	2.2059E+00
BI-207	163.	569.70	1.8095E+00
TL-208	515.	277.36	2.7881E+01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3398-DUP

Acquisition date : 15-JAN-1999 01:51:19

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	775.	46.50	0.0000E+00
BI-212	72.	727.18	2.2320E+01
PB-212	722.	238.63	4.2685E+00
BI-214	245.	609.31	5.1003E+00
PB-214	490.	351.92	2.2729E+01
RA-224	645.	240.99	4.5156E+01
RA-226	754.	186.10	4.4200E+01
AC-228	89.	911.21	7.4651E+00
TH-228	610.	84.37	1.5016E+02
TH-229	653.	88.47	7.3080E+00
U-232	484.	57.78	3.5603E+03
PA-233	486.	312.17	4.8400E+00
UTH-233	576.	245.34	1.5477E+03
PA-234M	45.	1001.03	2.4206E+02
TH-234	562.	63.29	1.1506E+02
U-235	742.	185.71	2.6661E+00
NP-237	668.	86.48	1.5309E+01
U-237	515.	59.54	1.7192E+01
NP-238	45.	984.45	5.6277E+00
NP-239	625.	106.12	6.2223E+00
PU-239	645.	129.30	1.9928E+04
AM-241	515.	59.54	1.6432E+01
AM-243	667.	74.67	4.3823E+00

*****

* 222-S Laboratory Counting Room 15-JAN-1999 07:05:37.64 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3388-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

En Barla

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3143.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:13.13 sec
 Dead Time: 0.1%

Verified by:

John Relyea 15 Jan 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 15-JAN-1999 04:34:56.06
 Decayed to: 15-JAN-1999 04:34:56.06
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.79*	88587	1.59	1323.75	1315	18	0.7		CS-137	33.4

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3388-SAM

Acquisition date : 15-JAN-1999 04:34:56

Total number of lines in spectrum	1	
Number of unidentified lines	0	
Number of lines tentatively identified by NID	1	100.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	3.338E+01	3.338E+01	0.023E+01	0.69	
Total Activity :			3.338E+01	3.338E+01			

Grand Total Activity :	3.338E+01	3.338E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3388-SAM

Acquisition date : 15-JAN-1999 04:34:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2974.	477.59	6.0549E-01
NA-24	14.	1368.55	1.2507E-02
K-40	374.	1460.75	5.4328E-01
AR-41	15.	1293.60	1.8446E-02
SC-46	60.	1120.55	1.9553E-02
CR-51	2334.	320.08	4.0440E-01
MN-54	62.	834.83	1.5399E-02
CO-56	41.	846.76	1.2866E-02
CO-58	56.	810.78	1.4445E-02
FE-59	51.	1099.25	3.1601E-02
CO-60	36.	1332.50	1.7824E-02
ZN-65	44.	1115.55	3.3200E-02
SE-75	2734.	264.66	6.5470E-02
KR-85	1526.	514.00	1.1008E+01
SR-85	1525.	514.01	4.9779E-02
Y-88	9.	1836.06	1.2392E-02
Y-91	28.	1204.67	4.8851E+00
NB-94	46.	871.09	1.3859E-02
NB-95	61.	765.78	1.4274E-02
ZR-95	57.	756.73	2.5025E-02
ZRNB-95	81.	724.18	7.1113E-02
RU-103	1850.	497.08	5.6251E-02
RURH-106	565.	621.93	6.9588E-01
AG-108m	73.	722.94	1.6168E-02
CD-109	2332.	88.03	9.3370E-01
AG-110M	1956.	657.76	7.0360E-02
SN-113	2499.	391.69	7.6656E-02
TE-123m	2750.	159.00	3.1474E-02
SB-124	591.	602.73	3.5134E-02
SB-125	3005.	427.89	1.9600E-01
TE-125m	2448.	109.27	9.4029E+00
I-129	2477.	39.60	0.0000E+00
I-131	2399.	364.48	5.6326E-02
XE-131m	2671.	163.93	1.3518E+00
BA-133	2335.	356.02	7.1148E-02
CS-134	565.	604.70	3.4511E-02
CS-136	49.	818.51	1.3649E-02
CS-138	17.	1435.86	5.8936E-02
CE-139	2701.	165.85	3.3452E-02
BA-140	1015.	537.31	1.6675E-01
LA-140	18.	1596.21	1.6262E-02
CE-141	2766.	145.44	5.3624E-02
CE-144	2675.	133.51	2.2840E-01
CEPR-144	2675.	133.51	4.5641E-01
EU-152	2711.	121.78	9.1948E-02
EU-154	40.	1274.51	5.2504E-02
EU-155	2390.	86.54	1.0715E-01
HF-181	2394.	482.18	7.0487E-02
TA-182	2133.	67.75	1.5218E-01
HG-203	2611.	279.20	4.7737E-02
BI-207	664.	569.70	3.5418E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3388-SAM

Acquisition date : 15-JAN-1999 04:34:56

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2608.	277.36	6.1260E-01
PB-210	2339.	46.50	0.0000E+00
BI-212	83.	727.18	2.3651E-01
PB-212	3037.	238.63	8.5750E-02
BI-214	703.	609.31	8.4202E-02
PB-214	2427.	351.92	4.9427E-01
RA-224	2923.	240.99	9.4045E-01
RA-226	3543.	186.10	9.3809E-01
AC-228	85.	911.21	7.2560E-02
TH-228	2339.	84.37	2.8795E+00
TH-229	2333.	88.47	1.3533E-01
U-232	2055.	57.78	7.1643E+01
PA-233	2394.	312.17	1.0491E-01
UTH-233	2944.	245.34	3.4182E+01
PA-234M	36.	1001.03	2.1686E+00
TH-234	2185.	63.29	2.2205E+00
U-235	3478.	185.71	5.6498E-02
NP-237	2390.	86.48	2.8381E-01
U-237	2042.	59.54	3.3483E-01
NP-238	32.	984.45	4.8254E-02
NP-239	2433.	106.12	1.2016E-01
PU-239	2694.	129.30	3.9851E+02
AM-241	2042.	59.54	3.2002E-01
AM-243	2354.	74.67	8.0702E-02

*****
 * 222-S Laboratory Counting Room 15-JAN-1999 09:40:30.72 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 27865
 Sample ID: S98T3388-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA3
 File Number: dka300:[spec.GEA3]3g3144.cnf
 Geometry: 42
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:12.64 sec
 Dead Time: 0.1%

Verified by:

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 15-JAN-1999 07:09:48.41
 Decayed to: 15-JAN-1999 07:09:48.41
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: EMB
 Background Subtract: DKA300:[SPEC.GEA3]3GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 24-SEP-1998 08:17:01.62
 Date of last efficiency calibration: 15-MAR-1994 10:28:40.20

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	661.78*	84846	1.60	1323.73	1315	17	0.7		CS-137	32.0
0	1427.41	16	1.11	2855.11	2848	11	87.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T3388-DUP

Acquisition date : 15-JAN-1999 07:09:48

Total number of lines in spectrum	2	
Number of unidentified lines	1	
Number of lines tentatively identified by NID	1	50.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CS-137	30.00Y	1.00	3.197E+01	3.197E+01	0.022E+01	0.70	
Total Activity :			3.197E+01	3.197E+01			

Grand Total Activity : 3.197E+01   3.197E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T3388-DUP

Acquisition date : 15-JAN-1999 07:09:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	2837.	477.59	5.9153E-01
NA-24	26.	1368.55	1.6587E-02
K-40	370.	1460.75	5.4026E-01
AR-41	30.	1293.60	2.4870E-02
SC-46	83.	1120.55	2.2766E-02
CR-51	2224.	320.08	3.9488E-01
MN-54	51.	834.83	1.4083E-02
CO-56	53.	846.76	1.4561E-02
CO-58	64.	810.78	1.5286E-02
FE-59	42.	1099.25	2.8736E-02
CO-60	23.	1332.50	1.4694E-02
ZN-65	41.	1115.55	3.2103E-02
SE-75	2569.	264.66	6.3479E-02
KR-85	1422.	514.00	1.0633E+01
SR-85	1422.	514.01	4.8090E-02
Y-88	6.	1836.06	1.0573E-02
Y-91	50.	1204.67	6.3903E+00
NB-94	47.	871.09	1.4057E-02
NB-95	64.	765.78	1.4571E-02
ZR-95	56.	756.73	2.4876E-02
ZRNB-95	77.	724.18	6.9399E-02
RU-103	1665.	497.08	5.3401E-02
RURH-106	567.	621.93	6.9701E-01
AG-108m	72.	722.94	1.6052E-02
CD-109	2284.	88.03	9.2414E-01
AG-110M	1789.	657.76	6.7341E-02
SN-113	2544.	391.69	7.7333E-02
TE-123m	2578.	159.00	3.0483E-02
SB-124	606.	602.73	3.5574E-02
SB-125	2838.	427.89	1.9052E-01
TE-125m	2307.	109.27	9.1298E+00
I-129	2404.	39.60	0.0000E+00
I-131	2365.	364.48	5.5931E-02
XE-131m	2618.	163.93	1.3384E+00
BA-133	2230.	356.02	6.9551E-02
CS-134	586.	604.70	3.5125E-02
CS-136	45.	818.51	1.3142E-02
CS-138	14.	1435.86	5.4112E-02
CE-139	2563.	165.85	3.2594E-02
BA-140	976.	537.31	1.6363E-01
LA-140	16.	1596.21	1.5463E-02
CE-141	2571.	145.44	5.1722E-02
CE-144	2559.	133.51	2.2343E-01
CEPR-144	2559.	133.51	4.4647E-01
EU-152	2525.	121.78	8.8767E-02
EU-154	39.	1274.51	5.2000E-02
EU-155	2229.	86.54	1.0352E-01
HF-181	2338.	482.18	6.9667E-02
TA-182	2091.	67.75	1.5071E-01
HG-203	2404.	279.20	4.5822E-02
BI-207	608.	569.70	3.3925E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T3388-DUP

Acquisition date : 15-JAN-1999 07:09:48

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
TL-208	2475.	277.36	5.9696E-01
PB-210	2413.	46.50	0.0000E+00
BI-212	82.	727.18	2.3480E-01
PB-212	2898.	238.63	8.3785E-02
BI-214	665.	609.31	8.1968E-02
PB-214	2345.	351.92	4.8588E-01
RA-224	2782.	240.99	9.1770E-01
RA-226	3454.	186.10	9.2632E-01
AC-228	81.	911.21	7.0544E-02
TH-228	2266.	84.37	2.8349E+00
TH-229	2259.	88.47	1.3319E-01
U-232	2063.	57.78	7.1777E+01
PA-233	2281.	312.17	1.0242E-01
UTH-233	2828.	245.34	3.3514E+01
PA-234M	45.	1001.03	2.3794E+00
TH-234	2099.	63.29	2.1768E+00
U-235	3418.	185.71	5.6013E-02
NP-237	2231.	86.48	2.7431E-01
U-237	2094.	59.54	3.3904E-01
NP-238	42.	984.45	5.4506E-02
NP-239	2204.	106.12	1.1444E-01
PU-239	2625.	129.30	3.9345E+02
AM-241	2094.	59.54	3.2404E-01
AM-243	2263.	74.67	7.9132E-02

LABCORE Completed Worklist Report for Worklist# 28056

Analyst: scl

Instrument: GEA01

Book#: 28B57

Method: LA-548-121

Rev/Mod F-0

Worklist Comment: BY105 FIELD BLANK FOR @GEA-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@GEA-01 C060-02 LIQUID		4.400e-03	4.71e-03	107.045	% Recovery
1 STD	0		@GEA-01 C060-02E LIQUID		1	3.17	3.170	% Ct Error
1 STD	0		@GEA-01 CS13702 LIQUID		7.415e-03	7.68e-03	103.574	% Recovery
1 STD	0		@GEA-01 CS13702E LIQUID		1	2.76	2.760	% Ct Error
2 BLNK	0		@GEA-01 C060-02 LIQUID		1	<2.48e-5		uCi/mL
2 BLNK	0		@GEA-01 CS13702 LIQUID		1	<3.30e-5		uCi/mL
3 SAMPLE	S98T002391	0	@GEA-01 C060-02 LIQUID		N/A	<2.617e-05	2.62e-005	uCi/mL
3 SAMPLE	S98T002391	0	@GEA-01 C060-02E LIQUID		N/A	n/a		% Ct. Error
3 SAMPLE	S98T002391	0	@GEA-01 CS13702 LIQUID		N/A	<3.345e-05	3.34e-005	uCi/mL
3 SAMPLE	S98T002391	0	@GEA-01 CS13702E LIQUID		N/A	n/a		% Ct. Error
4 DUP	S98T002391	0	@GEA-01 C060-02 LIQUID		<2.62e-5	<2.84e-5		RPD
4 DUP	S98T002391	0	@GEA-01 C060-02E LIQUID		1	n/a		% Ct Error
4 DUP	S98T002391	0	@GEA-01 CS13702 LIQUID		<3.34e-5	<4.13e-5		RPD
4 DUP	S98T002391	0	@GEA-01 CS13702E LIQUID		1	n/a		% Ct Error

Final page for worklist# 28056

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

R Jones 2-2-99
Reviewer/Signature Date

LABCORE Data Entry Template for Worklist# 28056

Analyst: S. L. Instrument: GEA00 1 Book# 28 B57

Method: LA-548-121 Rev/Mod F-0

Worklist Comment: BY105 FIELD BLANK FOR @GEA-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@GEA-01	LIQUID		
2	BLNK				@GEA-01	LIQUID		
3	SAMPLE	S98T002391 0			@GEA-01	LIQUID	98000436	BY-105 (2)
					Analytes Requested: CO60-02 , CO60-02E, CS13702 , CS13702E			
4	DUP	S98T002391 0			@GEA-01	LIQUID		

Final page for worklist # 28056

Sue Lai 2-1-99
Analyst Signature Date

[Signature] 2/2/99
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

*****

* 222-S Laboratory Counting Room 1-FEB-1999 15:23:21.54 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 28056
 Sample ID: WL28056-STD
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

Raymond

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEAL
 File Number: dka300:[spec.GEAL]1g1031.cnf
 Geometry: 41
 Count Time: 0 00:50:00.00 sec
 Real Time: 0 00:50:14.78 sec
 Dead Time: 0.5%

Verified by:

John Rehyea 2 Feb 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 1-FEB-1999 14:32:36.44
 Decayed to: 1-FEB-1999 14:32:36.44
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEAL]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	32.01	1020	1.05	64.13	60	9 13.0			
0	36.29	291	0.86	72.71	69	9 37.8			
0	59.68	378	0.88	119.47	116	8 27.2		AM-241	0.262
0	661.78*	6341	1.43	1323.66	1316	14 2.8		CS-137	7.68
0	1173.26	2743	1.79	2346.81	2337	18 4.6		CO-60	4.72
0	1332.54	2467	1.99	2665.47	2655	19 4.4		CO-60	4.71
0	1764.64*	6	1.82	3530.04	3522	14255.6			

Summary of Nuclide Activity

Page : 2

Sample ID : WL28056-STD

Acquisition date : 1-FEB-1999 14:32:36

Total number of lines in spectrum	7	
Number of unidentified lines	2	
Number of lines tentatively identified by NID	5	71.43%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
CO-60	5.27Y	1.00	4.713E+00	4.713E+00	0.149E+00	3.17	
CS-137	30.00Y	1.00	7.678E+00	7.678E+00	0.212E+00	2.76	
AM-241	432.20Y*****		2.619E-01	2.619E-01	0.713E-01	27.25	
Total Activity :			1.265E+01	1.265E+01			

Grand Total Activity :	1.265E+01	1.265E+01

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL28056-STD

Acquisition date : 1-FEB-1999 14:32:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	416.	477.59	7.1798E-01
NA-24	25.	1368.55	5.1900E-02
K-40	575.	1460.75	2.1928E+00
AR-41	58.	1293.60	8.3262E-02
SC-46	187.	1120.55	1.0972E-01
CR-51	457.	320.08	5.3962E-01
MN-54	195.	834.83	8.6514E-02
CO-56	197.	846.76	8.8037E-02
CO-58	206.	810.78	8.6968E-02
FE-59	198.	1099.25	1.9630E-01
ZN-65	181.	1115.55	2.1240E-01
SE-75	533.	264.66	8.5595E-02
KR-85	284.	514.00	1.5244E+01
SR-85	284.	514.01	6.8911E-02
Y-88	7.	1836.06	3.6148E-02
Y-91	83.	1204.67	2.6337E+01
NB-94	209.	871.09	9.2902E-02
NB-95	170.	765.78	7.5035E-02
ZR-95	192.	756.73	1.4386E-01
ZRNB-95	176.	724.18	3.3097E-01
RU-103	309.	497.08	7.3380E-02
RURH-106	192.	621.93	1.3193E+00
AG-108m	168.	722.94	7.7602E-02
CD-109	614.	88.03	8.2212E-01
AG-110M	342.	657.76	9.6180E-02
SN-113	398.	391.69	9.4873E-02
TE-123m	666.	159.00	4.6501E-02
SB-124	180.	602.73	6.2921E-02
SB-125	399.	427.89	2.2390E-01
TE-125m	673.	109.27	1.1670E+01
I-129	454.	39.60	3.8916E-01
I-131	391.	364.48	6.9583E-02
XE-131m	717.	163.93	2.1051E+00
BA-133	429.	356.02	9.3161E-02
CS-134	189.	604.70	6.4840E-02
CS-136	177.	818.51	8.1435E-02
CS-138	22.	1435.86	1.0688E-01
CE-139	705.	165.85	5.1544E-02
BA-140	246.	537.31	2.6371E-01
LA-140	16.	1596.21	4.9163E-02
CE-141	697.	145.44	7.8871E-02
CE-144	748.	133.51	3.3963E-01
CEPR-144	748.	133.51	6.7871E-01
EU-152	705.	121.78	1.2363E-01
EU-154	50.	1274.51	1.9073E-01
EU-155	622.	86.54	9.0752E-02
HF-181	370.	482.18	8.8072E-02
TA-182	498.	67.75	6.3581E-02
HG-203	512.	279.20	6.2956E-02
BI-207	259.	569.70	7.1115E-02
TL-208	558.	277.36	8.4269E-01

4649

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL28056-STD

Acquisition date : 1-FEB-1999 14:32:36

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	530.	46.50	7.2109E-01
BI-212	194.	727.18	1.1433E+00
PB-212	657.	238.63	1.1724E-01
BI-214	320.	609.31	1.8320E-01
PB-214	531.	351.92	3.1698E-01
RA-224	592.	240.99	1.2373E+00
RA-226	727.	186.10	1.2846E+00
AC-228	317.	911.21	4.4391E-01
TH-228	642.	84.37	2.3774E+00
TH-229	609.	88.47	1.1978E-01
U-232	580.	57.78	1.4339E+01
PA-233	446.	312.17	1.3640E-01
UTH-233	511.	245.34	4.1991E+01
PA-234M	187.	1001.03	1.5297E+01
TH-234	547.	63.29	7.2457E-01
U-235	741.	185.71	7.8826E-02
NP-237	626.	86.48	2.4067E-01
U-237	878.	59.54	1.0150E-01
NP-238	189.	984.45	3.5638E-01
NP-239	624.	106.12	1.3779E-01
PU-239	739.	129.30	5.7559E+02
AM-243	690.	74.67	5.1296E-02

*****
 * 222-S Laboratory Counting Room 1-FEB-1999 18:07:30.62 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 28056
 Sample ID: WL28056-BLK
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

J. U. Lake

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1032.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:31.74 sec
 Dead Time: 0.4%

Verified by:

John Relyea 27 Feb 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 1-FEB-1999 15:36:28.92
 Decayed to: 1-FEB-1999 15:36:28.92
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: VR
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uCi/L
0	12.93	138	1.40	25.98	24	5 66.2			
0	16.20	294	1.21	32.53	29	8 44.0			
0	934.23	54	1.44	1868.65	1863	10 78.1			
0	1730.31	16	0.94	3461.34	3457	8 96.7	2.15E+00		

Summary of Nuclide Activity

Page : 2

Sample ID : WL28056-BLK

Acquisition date : 1-FEB-1999 15:36:28

Total number of lines in spectrum

4

Number of unidentified lines

4

Number of lines tentatively identified by NID

0

0.00%

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : WL28056-BLK

Acquisition date : 1-FEB-1999 15:36:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	289.	477.59	2.0079E-01
NA-24	62.	1368.55	2.7043E-02
K-40	1786.	1460.75	1.2747E+00
AR-41	82.	1293.60	4.3450E-02
SC-46	268.	1120.55	4.3465E-02
CR-51	615.	320.08	2.0812E-01
MN-54	163.	834.83	2.6470E-02
CO-56	142.	846.76	2.5076E-02
CO-58	178.	810.78	2.7010E-02
FE-59	129.	1099.25	5.3344E-02
CO-60	61.	1332.50	2.4782E-02
ZN-65	149.	1115.55	6.4499E-02
SE-75	788.	264.66	3.4533E-02
KR-85	392.	514.00	5.9384E+00
SR-85	391.	514.01	2.6813E-02
Y-88	26.	1836.06	2.1226E-02
Y-91	162.	1204.67	1.2097E+01
NB-94	173.	871.09	2.8326E-02
NB-95	194.	765.78	2.6607E-02
ZR-95	184.	756.73	4.7054E-02
ZRNB-95	182.	724.18	1.1239E-01
RU-103	310.	497.08	2.4504E-02
RURH-106	224.	621.93	4.7375E-01
AG-108m	187.	722.94	2.7236E-02
CD-109	1225.	88.03	3.8431E-01
AG-110M	220.	657.76	2.5927E-02
SN-113	414.	391.69	3.2252E-02
TE-123m	1330.	159.00	2.1768E-02
SB-124	252.	602.73	2.4692E-02
SB-125	374.	427.89	7.2359E-02
TE-125m	1349.	109.27	5.4720E+00
I-129	769.	39.60	1.6785E-01
I-131	496.	364.48	2.6136E-02
XE-131m	1306.	163.93	9.4399E-01
BA-133	494.	356.02	3.3273E-02
CS-134	249.	604.70	2.4678E-02
CS-136	147.	818.51	2.4842E-02
CS-137	289.	661.66	3.2971E-02
CS-138	50.	1435.86	1.0564E-01
CE-139	1274.	165.85	2.2974E-02
BA-140	273.	537.31	9.2714E-02
LA-140	38.	1596.21	2.4626E-02
CE-141	1417.	145.44	3.7280E-02
CE-144	1434.	133.51	1.5592E-01
CEPR-144	1434.	133.51	3.1154E-01
EU-152	1404.	121.78	5.7783E-02
EU-154	102.	1274.51	8.8700E-02
EU-155	1227.	86.54	4.2196E-02
HF-181	329.	482.18	2.7729E-02
TA-182	923.	67.75	2.8673E-02
HG-203	712.	279.20	2.4662E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : WL28056-BLK

Acquisition date : 1-FEB-1999 15:36:28

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	251.	569.70	2.3355E-02
TL-208	749.	277.36	3.2437E-01
PB-210	903.	46.50	3.1287E-01
BI-212	246.	727.18	4.2726E-01
PB-212	1185.	238.63	5.2184E-02
BI-214	590.	609.31	8.2257E-02
PB-214	786.	351.92	2.8474E-01
RA-224	992.	240.99	5.3460E-01
RA-226	1256.	186.10	5.6012E-01
AC-228	367.	911.21	1.5901E-01
TH-228	1333.	84.37	1.1340E+00
TH-229	1225.	88.47	5.6216E-02
U-232	1076.	57.78	6.4683E+00
PA-233	567.	312.17	5.1139E-02
UTH-233	819.	245.34	1.7624E+01
PA-234M	160.	1001.03	4.7319E+00
TH-234	1076.	63.29	3.3644E-01
U-235	1276.	185.71	3.4309E-02
NP-237	1239.	86.48	1.1214E-01
U-237	1017.	59.54	3.6499E-02
NP-238	132.	984.45	1.0117E-01
NP-239	1318.	106.12	6.6980E-02
PU-239	1470.	129.30	2.6896E+02
AM-241	1017.	59.54	3.4887E-02
AM-243	1355.	74.67	2.3803E-02

*****

* 222-S Laboratory Counting Room 1-FEB-1999 20:48:21.42 *

*****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<<

Worklist #: 28056
 Sample ID: S98T2391-SAM
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

SLH

>>>>>>>> COUNT INFORMATION <<<<<<<<<<

Detector ID: GEA1
 File Number: dka300:[spec.GEA1]1g1033.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:31.05 sec
 Dead Time: 0.3%

Verified by:

John Relyea 2 Feb 99

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<<

Sample Count Time: 1-FEB-1999 18:17:20.74
 Decayed to: 1-FEB-1999 18:17:20.74
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEA1]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw %Err	Fit	Nuclides	Activity uci/L
0	12.84	143	0.96	25.81	24	5 63.2			
0	15.94	261	1.32	32.00	29	8 49.5			

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2391-SAM

Acquisition date : 1-FEB-1999 18:17:20

Total number of lines in spectrum

2

Number of unidentified lines

2

Number of lines tentatively identified by NID

0

0.00%

**** There are no nuclides meeting summary criteria ****

Flags: "K" = Keyline not found

"M" = Manually accepted

"E" = Manually edited

"A" = Nuclide specific abn. limit

Minimum Detectable Activity Report
Sample ID : S98T2391-SAM

HNF-1663 REV.0
Page : 3
Acquisition date : 1-FEB-1999 18:17:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	320.	477.59	2.1101E-01
NA-24	55.	1368.55	2.5474E-02
K-40	1757.	1460.75	1.2643E+00
AR-41	79.	1293.60	4.2596E-02
SC-46	281.	1120.55	4.4494E-02
CR-51	582.	320.08	2.0254E-01
MN-54	177.	834.83	2.7485E-02
CO-56	161.	846.76	2.6610E-02
CO-58	156.	810.78	2.5359E-02
FE-59	132.	1099.25	5.3917E-02
CO-60	68.	1332.50	2.6172E-02
ZN-65	128.	1115.55	5.9965E-02
SE-75	766.	264.66	3.4057E-02
KR-85	369.	514.00	5.7628E+00
SR-85	368.	514.01	2.6033E-02
Y-88	20.	1836.06	1.9138E-02
Y-91	158.	1204.67	1.1961E+01
NB-94	162.	871.09	2.7433E-02
NB-95	223.	765.78	2.8493E-02
ZR-95	203.	756.73	4.9321E-02
ZRNB-95	188.	724.18	1.1419E-01
RU-103	313.	497.08	2.4616E-02
RURH-106	255.	621.93	5.0338E-01
AG-108m	189.	722.94	2.7391E-02
CD-109	1210.	88.03	3.8203E-01
AG-110M	199.	657.76	2.4683E-02
SN-113	434.	391.69	3.3007E-02
TE-123m	1368.	159.00	2.2072E-02
SB-124	249.	602.73	2.4538E-02
SB-125	396.	427.89	7.4376E-02
TE-125m	1284.	109.27	5.3415E+00
I-129	762.	39.60	1.6715E-01
I-131	486.	364.48	2.5880E-02
XE-131m	1321.	163.93	9.4919E-01
BA-133	510.	356.02	3.3815E-02
CS-134	226.	604.70	2.3546E-02
CS-136	149.	818.51	2.5010E-02
CS-137	297.	661.66	3.3454E-02
CS-138	45.	1435.86	1.0105E-01
CE-139	1300.	165.85	2.3205E-02
BA-140	259.	537.31	9.0277E-02
LA-140	35.	1596.21	2.3449E-02
CE-141	1457.	145.44	3.7788E-02
CE-144	1430.	133.51	1.5569E-01
CEPR-144	1430.	133.51	3.1109E-01
EU-152	1419.	121.78	5.8090E-02
EU-154	98.	1274.51	8.7073E-02
EU-155	1289.	86.54	4.3232E-02
HF-181	301.	482.18	2.6552E-02
TA-182	956.	67.75	2.9173E-02
HG-203	756.	279.20	2.5389E-02

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2391-SAM

Acquisition date : 1-FEB-1999 18:17:20

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BI-207	265.	569.70	2.3975E-02
TL-208	752.	277.36	3.2489E-01
PB-210	901.	46.50	3.1253E-01
BI-212	247.	727.18	4.2732E-01
PB-212	1252.	238.63	5.3619E-02
BI-214	575.	609.31	8.1198E-02
PB-214	829.	351.92	2.9222E-01
RA-224	1056.	240.99	5.5109E-01
RA-226	1264.	186.10	5.6177E-01
AC-228	330.	911.21	1.5102E-01
TH-228	1294.	84.37	1.1176E+00
TH-229	1208.	88.47	5.5838E-02
U-232	1077.	57.78	6.4713E+00
PA-233	611.	312.17	5.3043E-02
UTH-233	839.	245.34	1.7839E+01
PA-234M	151.	1001.03	4.6016E+00
TH-234	1055.	63.29	3.3309E-01
U-235	1266.	185.71	3.4169E-02
NP-237	1298.	86.48	1.1474E-01
U-237	1051.	59.54	3.7087E-02
NP-238	140.	984.45	1.0427E-01
NP-239	1293.	106.12	6.6339E-02
PU-239	1439.	129.30	2.6613E+02
AM-241	1051.	59.54	3.5450E-02
AM-243	1296.	74.67	2.3287E-02

*****
 * 222-S Laboratory Counting Room 1-FEB-1999 23:22:41.40 *
 *****

>>>>>>>> SAMPLE INFORMATION <<<<<<<<<

Worklist #: 28056
 Sample ID: S98T2391-DUP
 Sample Size: 1.00000E-03 L
 Dilution Factor: 1.00000E+00

Removed by:

N. Miller

>>>>>>>> COUNT INFORMATION <<<<<<<<<

Detector ID: GEAL
 File Number: dka300:[spec.GEAL]1g1034.cnf
 Geometry: 41
 Count Time: 0 02:30:00.00 sec
 Real Time: 0 02:30:31.97 sec
 Dead Time: 0.4%

Verified by:

John Redgen 2 Feb 89

>>>>>>>> ANALYSIS INFORMATION <<<<<<<<<

Sample Count Time: 1-FEB-1999 20:51:37.22
 Decayed to: 1-FEB-1999 20:51:37.22
 Standard Deviations: 2
 Analysis Library: ENVGEA
 Analyst: SLH2
 Background Subtract: DKA300:[SPEC.GEAL]1GBACK

>>>>>>>> CALIBRATION INFORMATION <<<<<<<<<

Date of last energy calibration: 23-SEP-1998 14:57:30.35
 Date of last efficiency calibration: 20-MAR-1996 09:10:52.48

*****

Post-NID Peak Search Report

It	Energy	Area	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides	Activity uCi/L
0	16.13	284	1.25	32.39	29	8	46.5			
0	661.96*	102	4.42	1324.02	1317	18	92.5	2.39E+00	CS-137	4.128E-02
0	935.11	87	3.23	1870.39	1864	13	55.7			
0	970.36	85	1.54	1940.91	1932	18	99.7			
0	1460.89*	58	1.94	2922.27	2914	19	173.8		K-40	0.373

Summary of Nuclide Activity

Page : 2

Sample ID : S98T2391-DUP

Acquisition date : 1-FEB-1999 20:51:37

Total number of lines in spectrum 5
 Number of unidentified lines 2
 Number of lines tentatively identified by NID 3 60.00%

Nuclide Type :

Nuclide	Hlife	Decay	Wtd Mean Uncorrected uCi/L	Wtd Mean Decay Corr uCi/L	Decay Corr 2-Sigma Error	2-Sigma %Error	Flags
~~K-40~~	~~1.28E+09Y~~	~~1.00~~	~~3.733E-01~~	~~3.733E-01~~	~~6.487E-01~~	~~173.78~~	~~JP~~
~~CS-137~~	~~30.00Y~~	~~1.00~~	~~4.128E-02~~	~~4.128E-02~~	~~3.820E-02~~	~~92.53~~	~~JP~~
Total Activity :			4.146E-01	4.146E-01			

JP
2/2/99

Grand Total Activity : 4.146E-01 4.146E-01

Flags: "K" = Keyline not found
 "E" = Manually edited

"M" = Manually accepted
 "A" = Nuclide specific abn. limit

Minimum Detectable Activity Report

Page : 3

Sample ID : S98T2391-DUP

Acquisition date : 1-FEB-1999 20:51:37

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
BE-7	353.	477.59	2.2104E-01
NA-24	54.	1368.55	2.5423E-02
AR-41	85.	1293.60	4.4247E-02
SC-46	264.	1120.55	4.3177E-02
CR-51	608.	320.08	2.0691E-01
MN-54	172.	834.83	2.7163E-02
CO-56	174.	846.76	2.7642E-02
CO-58	187.	810.78	2.7675E-02
FE-59	128.	1099.25	5.3202E-02
CO-60	81.	1332.50	2.8404E-02
ZN-65	115.	1115.55	5.6990E-02
SE-75	770.	264.66	3.4147E-02
KR-85	372.	514.00	5.7896E+00
SR-85	372.	514.01	2.6170E-02
Y-88	29.	1836.06	2.2451E-02
Y-91	120.	1204.67	1.0498E+01
NB-94	159.	871.09	2.7149E-02
NB-95	196.	765.78	2.6746E-02
ZR-95	210.	756.73	5.0117E-02
ZRNB-95	208.	724.18	1.1958E-01
RU-103	317.	497.08	2.4773E-02
RURH-106	228.	621.93	4.7712E-01
AG-108m	198.	722.94	2.8010E-02
CD-109	1187.	88.03	3.7840E-01
AG-110M	207.	657.76	2.5171E-02
SN-113	423.	391.69	3.2591E-02
TE-123m	1395.	159.00	2.2291E-02
SB-124	266.	602.73	2.5345E-02
SB-125	371.	427.89	7.2102E-02
TE-125m	1365.	109.27	5.5033E+00
I-129	765.	39.60	1.6741E-01
I-131	499.	364.48	2.6213E-02
XE-131m	1328.	163.93	9.5194E-01
BA-133	492.	356.02	3.3216E-02
CS-134	252.	604.70	2.4807E-02
CS-136	169.	818.51	2.6589E-02
CS-138	47.	1435.86	1.0328E-01
CE-139	1285.	165.85	2.3079E-02
BA-140	291.	537.31	9.5488E-02
LA-140	39.	1596.21	2.4918E-02
CE-141	1346.	145.44	3.6350E-02
CE-144	1423.	133.51	1.5533E-01
CEPR-144	1423.	133.51	3.1036E-01
EU-152	1409.	121.78	5.7888E-02
EU-154	86.	1274.51	8.1882E-02
EU-155	1270.	86.54	4.2929E-02
HF-181	322.	482.18	2.7446E-02
TA-182	1031.	67.75	3.0276E-02
HG-203	753.	279.20	2.5346E-02
BI-207	282.	569.70	2.4706E-02
TL-208	767.	277.36	3.2806E-01

Minimum Detectable Activity Report (continued)

Page : 4

Sample ID : S98T2391-DUP

Acquisition date : 1-FEB-1999 20:51:37

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/L)
PB-210	898.	46.50	3.1207E-01
BI-212	228.	727.18	4.1131E-01
PB-212	1196.	238.63	5.2435E-02
BI-214	599.	609.31	8.2865E-02
PB-214	722.	351.92	2.7311E-01
RA-224	987.	240.99	5.3313E-01
RA-226	1291.	186.10	5.6754E-01
AC-228	388.	911.21	1.6326E-01
TH-228	1291.	84.37	1.1162E+00
TH-229	1158.	88.47	5.4701E-02
U-232	1073.	57.78	6.4600E+00
PA-233	603.	312.17	5.2672E-02
UTH-233	838.	245.34	1.7829E+01
PA-234M	133.	1001.03	4.3300E+00
TH-234	1129.	63.29	3.4440E-01
U-235	1287.	185.71	3.4448E-02
NP-237	1275.	86.48	1.1375E-01
U-237	1010.	59.54	3.6371E-02
NP-238	134.	984.45	1.0216E-01
NP-239	1305.	106.12	6.6657E-02
PU-239	1404.	129.30	2.6293E+02
AM-241	1010.	59.54	3.4765E-02
AM-243	1317.	74.67	2.3471E-02

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 16 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 4463 THRU 4822)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	STRONTIUM ANALYSIS, SR90 WORKLIST		

TABLE OF CONTENTS

Sec 1
of 7

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788	(2589, 2590)	391
DSC Worklist # 26789	(2591, 2592)	398
DSC Worklist # 26790	(2593, 2594)	404
DSC Worklist # 26791	(2711, 2737)	410
DSC Worklist # 26792	(2840, 2841)	416
DSC Worklist # 26793	(2910, 2911)	422
DSC Worklist # 26794	(2912, 2913)	428
DSC Worklist # 26795	(2914, 2980)	434
DSC Worklist # 26796	(2981, 2982, 2983)	440
DSC Worklist # 26852	(3002)	449
DSC Worklist # 26854	(3004, 3005)	456
DSC Worklist # 26916	(2957, 2958, 2959)	462
DSC Worklist # 26917	(2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918	(2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166	(2811)	465
DSC Worklist # 27173	(2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174	(2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230	(2811, 2812, 3004, 3005)	472
DSC Worklist # 27231	(2841)	473
DSC Worklist # 27633	(3383)	474
DSC Worklist # 27634	(3191, 3396)	479
DSC Worklist # 27643	(3386)	486
DSC Worklist # 27765	(3386, 3391, 3396)	491
DSC Worklist # 27811	(3383)	492
DSC Worklist # 28062	(2391)	493
DSC Worklist # 28888	(2391)	498
Thermogravimetric Analysis (TGA)		
TGA Worklist # 26765	(2806)	499
TGA Worklist # 26766	(2807, 2808)	503
TGA Worklist # 26767	(2809, 2810)	509
TGA Worklist # 26768	(2811, 2812)	515
TGA Worklist # 26769	(2813, 2938)	521
TGA Worklist # 26770	(2955, 2956)	527
TGA Worklist # 26771	(2957, 2958, 2959)	533
TGA Worklist # 26772	(2589, 2590)	541
TGA Worklist # 26773	(2591, 2592)	547
TGA Worklist # 26774	(2593, 2594)	553
TGA Worklist # 26775	(2711, 2737)	559
TGA Worklist # 26776	(2840, 2841)	565
TGA Worklist # 26777	(2910, 2911)	571
TGA Worklist # 26778	(2912, 2913)	581
TGA Worklist # 26779	(2914, 2980)	587
TGA Worklist # 26780	(2981, 2982, 2983)	593
TGA Worklist # 26851	(3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Gamma Energy Analysis (GEA)	
GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Strontium 90 Analysis (Sr90)	
Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Plutonium 239 Analysis (Pu239)	
Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26725

Analyst: scl Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY-105 FOR @SR90-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	0	@SR90-01 SR90-01 LIQUID		1.04E-03	1.06E-3	101.923 % Recovery	
1 STD	0	0	@SR90-01 SR90-01C LIQUID		100	9.24E+01	92.400 % Recovery	
1 STD	0	0	@SR90-01 SR90-01E LIQUID		1.00	1.85E+00	1.850 % Ct. Error	
2 BLNK	0	0	@SR90-01 SR90-01 LIQUID		1	1.56E-3	0.002 uCi/mL	
2 BLNK	0	0	@SR90-01 SR90-01C LIQUID		100	9.50E+01	95.000 % Recovery	
2 BLNK	0	0	@SR90-01 SR90-01E LIQUID		1.00	9.00E+01	90.000 uCi/mL	
3 BLNK/BKG	0	0	@SR90-01 SR90-01 LIQUID		1	1.77E+00	1.770 BLNK/BKG	
4 SAMPLE	S98T002818	0	@SR90-01 SR90-01 LIQUID		N/A	1.68E-01	1.74E-003 uCi/mL	
4 SAMPLE	S98T002818	0	@SR90-01 SR90-01C LIQUID		N/A	9.36E+01	% Recovery	
4 SAMPLE	S98T002818	0	@SR90-01 SR90-01E LIQUID		N/A	4.67E+00	% Ct. Error	
5 DUP	S98T002818	0	@SR90-01 SR90-01 LIQUID		1.68E-1	1.68E-1	0.000 RPD	
5 DUP	S98T002818	0	@SR90-01 SR90-01C LIQUID		100	9.31E+01	93.100 % Recovery	
5 DUP	S98T002818	0	@SR90-01 SR90-01E LIQUID		1.00	4.67E+00	4.670 % Ct. Err	
6 SAMPLE	S98T002819	0	@SR90-01 SR90-01 LIQUID		N/A	1.32E-01	1.75E-003 uCi/mL	
6 SAMPLE	S98T002819	0	@SR90-01 SR90-01C LIQUID		N/A	9.29E+01	% Recovery	
6 SAMPLE	S98T002819	0	@SR90-01 SR90-01E LIQUID		N/A	5.29E+00	% Ct. Error	
7 DUP	S98T002819	0	@SR90-01 SR90-01 LIQUID		1.32E-1	1.47E-1	10.753 RPD	
7 DUP	S98T002819	0	@SR90-01 SR90-01C LIQUID		100	9.33E+01	93.300 % Recovery	
7 DUP	S98T002819	0	@SR90-01 SR90-01E LIQUID		1.00	4.98E+00	4.980 % Ct. Err	
8 SAMPLE	S98T002820	0	@SR90-01 SR90-01 LIQUID		N/A	8.35E-02	1.77E-003 uCi/mL	
8 SAMPLE	S98T002820	0	@SR90-01 SR90-01C LIQUID		N/A	9.18E+01	% Recovery	
8 SAMPLE	S98T002820	0	@SR90-01 SR90-01E LIQUID		N/A	6.74E+00	% Ct. Error	
9 DUP	S98T002820	0	@SR90-01 SR90-01 LIQUID		8.35E-2	8.16E-2	2.302 RPD	
9 DUP	S98T002820	0	@SR90-01 SR90-01C LIQUID		100	9.37E+01	93.700 % Recovery	
9 DUP	S98T002820	0	@SR90-01 SR90-01E LIQUID		1.00	6.75E+00	6.750 % Ct. Err	
10 SAMPLE	S98T002821	0	@SR90-01 SR90-01 LIQUID		N/A	2.40E-02	1.74E-003 uCi/mL	
10 SAMPLE	S98T002821	0	@SR90-01 SR90-01C LIQUID		N/A	9.32E+01	% Recovery	
10 SAMPLE	S98T002821	0	@SR90-01 SR90-01E LIQUID		N/A	1.32E+01	% Ct. Error	
11 DUP	S98T002821	0	@SR90-01 SR90-01 LIQUID		2.40E-2	2.41E-2	0.416 RPD	
11 DUP	S98T002821	0	@SR90-01 SR90-01C LIQUID		100	9.34E+01	93.400 % Recovery	
11 DUP	S98T002821	0	@SR90-01 SR90-01E LIQUID		1.00	1.31E+01	13.100 % Ct. Err	
12 SAMPLE	S98T002939	0	@SR90-01 SR90-01 LIQUID		N/A	1.62E-02	1.73E-003 uCi/mL	
12 SAMPLE	S98T002939	0	@SR90-01 SR90-01C LIQUID		N/A	9.36E+01	% Recovery	
12 SAMPLE	S98T002939	0	@SR90-01 SR90-01E LIQUID		N/A	1.65E+01	% Ct. Error	
13 DUP	S98T002939	0	@SR90-01 SR90-01 LIQUID		1.62E-2	1.63E-2	0.615 RPD	
13 DUP	S98T002939	0	@SR90-01 SR90-01C LIQUID		100	9.41E+01	94.100 % Recovery	
13 DUP	S98T002939	0	@SR90-01 SR90-01E LIQUID		1.00	1.64E+01	16.400 % Ct. Err	

Units shown for QC (BLK/BKG) may not reflect the actual units.

10/22/98 15:39
A-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26725

Analyst: SL Instrument: AB00 12 Book# 28857Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY-105 FOR @SR90-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@SR90-01	LIQUID		
2	BLNK				@SR90-01	LIQUID		
3	BLNK/BKG				@SR90-01	LIQUID		
4	SAMPLE	S98T002818 0			@SR90-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E							
5	DUP	S98T002818 0			@SR90-01	LIQUID		
6	SAMPLE	S98T002819 0			@SR90-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E							
7	DUP	S98T002819 0			@SR90-01	LIQUID		
8	SAMPLE	S98T002820 0			@SR90-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E							
9	DUP	S98T002820 0			@SR90-01	LIQUID		
10	SAMPLE	S98T002821 0			@SR90-01	LIQUID	98000437	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E							
11	DUP	S98T002821 0			@SR90-01	LIQUID		
12	SAMPLE	S98T002939 0			@SR90-01	LIQUID	98000490	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E							
13	DUP	S98T002939 0			@SR90-01	LIQUID		

Final page for worklist # 26725

See Lan 12-4-98
Analyst Signature Date

W Raymond 12-7-98
Analyst Signature Date
NEB 12/2/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
TYPE	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	11241	GROSS WEIGHT (W2)	7.1417
WorkList	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.0493
26725	BACKGROUND in cpm (BKG)	2.2	NET WEIGHT (W3)	0.0924
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	21.42
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR (DDF)	1		
LIQUID	SAMPLE COUNT RATE (Rs)	1121.90	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.89	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005332	TIME OF SEPARATION (ST)	13:45	Rmax	N/A
ReRun	DATE OF SEPARATION (SD)	12/03/98	DETECTION LIMIT (Ld)	1.88
0	TIME OF COUNT (TOC)	11:10	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0636E+00
Sample Prep	DATE OF COUNT (DOC)	12/04/98		
N/A	STANDARD BOOK #	28B57		
Sample	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0416E-03		
STD				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS < Lc$ AND $RS > 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
VAR	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / 64.2 * DT)))) * SS * REC * 2220000$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
SCL	NOTE: Expected weight = CVA * 0.1			
Data Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
12/07/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date	12/04/98 Sr-89/90 CONCENTRATION 1.06E-03 $\mu\text{Ci/mL}$			DETECTION LEVEL 1.78E-06 $\mu\text{Ci/L}$
Analysis Time				
08:40 AM	RELATIVE COUNTING ERROR	1.9%		
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY	92.4%		

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	7 Dec 98

STANDARD.WB1 REV 2.0

22010NML

4666

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					BLNK
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
BLNK	TOTAL COUNTS	(TC)	39	GROSS WEIGHT	(W2) 7.1764
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.0814
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0950
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 23.33
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1.70	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005332	TIME OF SEPARATION	(ST)	13:45	Rmax	N/A
Run	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	13:05	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.5603E+00
Sample Prep	DATE OF COUNT	(DOC)	12/04/98		
N/A					
Sample #					
BLNK					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date					
12/04/98	Sr-89/90 CONCENTRATION	1.56E-03	$\mu\text{Ci/mL}$	DETECTION LEVEL 1.72E-03 $\mu\text{Ci/mL}$	
Analysis Time					
08:40 AM	RELATIVE COUNTING ERROR	90.0%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	95.0%			

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist: <i>[Signature]</i>	SAC	Date:	<i>12 Dec 98</i>

BLANK.WB1 REV 2.0

22010NML

4667

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Any type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	1827	GROSS WEIGHT	(W2)	7.1020
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0084
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	0.0936
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	23.80
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	180.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005332	TIME OF SEPARATION	(ST)	13:45	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	13:33	Sr-89/90 CONC in µCi/L		1.6756E+02
Sample Prep	DATE OF COUNT	(DOC)	12/04/98			
N/A						
Sample #						
S98T002818						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Completed	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						DETECTION LEVEL
12/04/98	Sr-89/90 CONCENTRATION		1.68E-01	µCi/mL		
Analysis Time						
08:40 AM	RELATIVE COUNTING ERROR		4.7%		1.74E-03	
Sample Point						µCi/mL
BY-105 (2)	PERCENT CARRIER RECOVERY		93.6%			

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: <i>7 Dec 98</i>
SAMPLE.WB1 REV 2.0 22010NML			

4668

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	1827	GROSS WEIGHT	(W2)	1.000
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.8873
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	7.7942
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	0.0931
@SR90-01	DILUTION FACTOR	DF	101			24.00
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	180.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005332	TIME OF SEPARATION	(ST)	13:45	Rmax		N/A
Repun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	13:45	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.6821E+02
Sample Prep	DATE OF COUNT	(DOC)	12/04/98			
N/A						
Sample #						
S98T002818						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						
12/04/98	Sr-89/90 CONCENTRATION		1.68E-01	$\mu\text{Ci/mL}$		DETECTION LEVEL . 1.75E-03 $\mu\text{Ci/mL}$
Analysis Time						
08:40 AM	RELATIVE COUNTING ERROR		4.7%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		93.1%			

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	<i>DA Cath</i>	Date:	7 Dec 98
SAMPLE.WB1 REV 2.0	22010NML		

4669

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)						SAMPLE
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000	
SAMPLE	TOTAL COUNTS (TC)	1439	GROSS WEIGHT (W2)		7.8774	
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.7845	
26725	BACKGROUND in cpm (BKG)	2.2	NET WEIGHT (W3)		0.0929	
Test Code	SAMPLE VOLUME in mL (SS)	0.100	DELTA TIME (HOURS) (DT)		24.10	
@SR90-01	DILUTION FACTOR (DF)	101				
Matrix	DIGEST DILUTION FACTOR (DDF)	1				
LIQUID	SAMPLE COUNT RATE (Rs)	141.70	SR-90 EFFICIENCY FACTOR (C1)		0.4180	
Batch Number	CRITICAL LEVEL (Lc)	0.89	Y-90 EFFICIENCY FACTOR (C2)		0.4660	
98005332	TIME OF SEPARATION (ST)	13:50	Rmax		N/A	
Rerun	DATE OF SEPARATION (SD)	12/03/98	DETECTION LIMIT (Ld)		1.88	
0	TIME OF COUNT (TOC)	13:56	Sr-89/90 CONC in µCi/L			1.3224E+02
Sample Prep	DATE OF COUNT (DOC)	12/04/98				
N/A						
Sample #						
S98T002819						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						DETECTION
12/04/98	Sr-89/90 CONCENTRATION	1.32E-01	µCi/mL			LEVEL
Analysis Time						
08:40 AM	RELATIVE COUNTING ERROR	5.3%				1.75E-03
Sample Point						µCi/mL
BY-105 (2)	PERCENT CARRIER RECOVERY	92.9%				

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	<i>[Signature]</i> SAC	Date:	<i>7 Dec 98</i>

SAMPLE.WB1 REV 2.0

22010NML

4C70

12/07/98

09:47:29

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
TYPE	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	1611	GROSS WEIGHT	(W2) 7.4010
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3077
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0933
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 24.28
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	158.90	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005332	TIME OF SEPARATION	(ST)	13:50	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	14:07	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.4746E+02
Sample Prep	DATE OF COUNT	(DOC)	12/04/98		
N/A					
Sample #					
S98T002819					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} \geq 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} \cdot \text{DF} \cdot \text{DDF} / ((\text{C1} + \text{C2} \cdot (1 - e^{-(\ln(2) \cdot 64.2 \cdot \text{DT}))}) \cdot \text{SS} \cdot \text{REC} \cdot 2220000)$				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) \cdot 24) + (\text{TOC} - \text{ST}) / 100$				
12/07/98					
Analysis Date					
12/04/98	Sr-89/90 CONCENTRATION	1.47E-01	$\mu\text{Ci/mL}$	DETECTION LEVEL 1.74E-03 $\mu\text{Ci/mL}$	
Analysis Time					
08:40 AM	RELATIVE COUNTING ERROR	5.0%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	93.3%			

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	7 Dec 98

SAMPLE.WB1 REV 2.0

22010NML

4671

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	909	GROSS WEIGHT	(W2)	7.4366
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3448
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	0.0918
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	24.47
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	88.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005332	TIME OF SEPARATION	(ST)	13:50	Rmax		N/A
SR90-01	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	14:18	Sr-89/90 CONC in $\mu\text{Ci/L}$		8.3544E+01
Sample Prep	DATE OF COUNT	(DOC)	12/04/98			
N/A						
Sample #						
S98T002820						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						DETECTION LEVEL 1.77E-03 $\mu\text{Ci/mL}$
12/04/98	Sr-89/90 CONCENTRATION	8.35E-02	$\mu\text{Ci/mL}$			
Analysis Time						
08:40 AM	RELATIVE COUNTING ERROR	6.7%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	91.8%				

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: <i>12 Dec 98</i>
SAMPLE.WB1 REV 2.0 22010NML			

4672

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP	
DUP	TOTAL COUNTS	(TC)	907	GROSS WEIGHT	(W2)	1.000	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.1336	
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	7.0399	
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	0.0937	
@SR90-01	DILUTION FACTOR	DF	101			24.63	
Matrix	DIGEST DILUTION FACTOR	(DDF)	1				
LIQUID	SAMPLE COUNT RATE	(Rs)	88.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005332	TIME OF SEPARATION	(ST)	13:50	Rmax		N/A	
Rerun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld)	1.88	
0	TIME OF COUNT	(TOC)	14:28	Sr-89/90 CONC in $\mu\text{Ci/L}$		8.1565E+01	
Sample Prep	DATE OF COUNT	(DOC)	12/04/98				
N/A							
Sample #							
S98T002820							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$						
Prepared By	$\text{RS} * \text{DF} * \text{DDF} / ((\text{C1} + \text{C2} * (1 - e^{-(\ln(2) * \text{DT})})) * \text{SS} * \text{REC} * 2220000)$						
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$						
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$						
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$						
12/07/98							
Analysis Date						DETECTION LEVEL 1.73E-03 $\mu\text{Ci/mL}$	
12/04/98	Sr-89/90 CONCENTRATION		8.16E-02	$\mu\text{Ci/mL}$			
Analysis Time							
08:40 AM	RELATIVE COUNTING ERROR		6.8%				
Sample Point							
BY-105 (2)	PERCENT CARRIER RECOVERY		93.7%				

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	<i>[Signature]</i>	Date:	7 Dec 98
SAMPLE.WB1 REV 2.0	22010NML		

4673

WORKBOOK PAGE: SAM10

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	281	GROSS WEIGHT (W2)		7.1195
WorkList	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.0263
26725	BACKGROUND in cpm (BKG)	2.2	NET WEIGHT (W3)		0.0932
Test Code	SAMPLE VOLUME in mL (SS)	0.100	DELTA TIME (HOURS) (DT)		24.82
@SR90-01	DILUTION FACTOR (DF)	101			
Main	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	25.90	SR-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.89	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98005332	TIME OF SEPARATION (ST)	13:55	Rmax		N/A
Refun	DATE OF SEPARATION (SD)	12/03/98	DETECTION LIMIT (Ld)		1.88
0	TIME OF COUNT (TOC)	14:44	Sr-89/90 CONC in $\mu\text{Ci/L}$ 2.3966E+01		
Sample Prep	DATE OF COUNT (DOC)	12/04/98			
N/A					
Sample					
S98T002821					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date	Sr-89/90 CONCENTRATION 2.40E-02 $\mu\text{Ci/mL}$				DETECTION LEVEL 1.74E-03 $\mu\text{Ci/mL}$
12/04/98					
Analysis Time					
08:40 AM	RELATIVE COUNTING ERROR 13.2%				
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY 93.2%				

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	7 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4674

WORKBOOK PAGE: DUP11

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	283	GROSS WEIGHT	(W2)	1.000
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.1371
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	7.0437
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	0.0934
@SR90-01	DILUTION FACTOR	DF	101			25.02
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	26.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005332	TIME OF SEPARATION	(ST)	13:55	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	14:56	Sr-89/90 CONC in $\mu\text{Ci/L}$		2.4065E+01
Sample Prep	DATE OF COUNT	(DOC)	12/04/98			
N/A						
Sample #						
S98T002821						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						
12/04/98	Sr-89/90 CONCENTRATION		2.41E-02	$\mu\text{Ci/mL}$		DETECTION LEVEL 1.73E-03 $\mu\text{Ci/mL}$
08:40 AM	RELATIVE COUNTING ERROR		13.1%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		93.4%			

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:		SAC	Date:
SAMPLE.WB1 REV 2.0	22010NML		7 Dec 98

4675

WORKBOOK PAGE: SAM12

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS (TC)		198	GROSS WEIGHT (W2)	7.1112
Work List	COUNT TIME in MINUTES (CT)		10	TARE WEIGHT (W1)	7.0176
26725	BACKGROUND in cpm (BKG)		2.2	NET WEIGHT (W3)	0.0936
Test Code	SAMPLE VOLUME in mL (SS)		0.100	DELTA TIME (HOURS) (DT)	25.20
@SR90-01	DILUTION FACTOR (DF)		101		
Matrix	DIGEST DILUTION FACTOR (DDF)		1		
LIQUID	SAMPLE COUNT RATE (Rs)		17.60	SR-90 EFFICIENCY FACTOR (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)		0.89	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005332	TIME OF SEPARATION (ST)		13:55	Rmax	N/A
ReRun	DATE OF SEPARATION (SD)		12/03/98	DETECTION LIMIT (Ld)	1.88
0	TIME OF COUNT (TOC)		15:07	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.6171E+01
Sample Prep	DATE OF COUNT (DOC)		12/04/98		
N/A					
Sample #					
S98T002939					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date					
12/04/98	Sr-89/90 CONCENTRATION		1.62E-02	$\mu\text{Ci/mL}$	DETECTION LEVEL 1.73E-03 $\mu\text{Ci/mL}$
Analysis Time					
08:40 AM	RELATIVE COUNTING ERROR		16.5%		
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY		93.6%		

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	2 Dec 98

SAMPLE WB1 REV 2.0

22010NML

4676

WORKBOOK PAGE: DUP13

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	201	GROSS WEIGHT	(W2) 7.1379
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.0438
26725	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0941
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 25.40
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	17.90	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005332	TIME OF SEPARATION	(ST)	13:55	Rmax	N/A
Retun	DATE OF SEPARATION	(SD)	12/03/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	15:19	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.6336E+01
Sample Prep	DATE OF COUNT	(DOC)	12/04/98		
N/A					
Sample #					
S98T002939					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} \cdot \text{DF} \cdot \text{DDF} / ((\text{C1} + \text{C2} \cdot (1 - e \text{ to the power of } ((-\text{natural log } 2) / (64.2 \cdot \text{DT})))) \cdot \text{SS} \cdot \text{REC} \cdot 2220000)$				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})$) $\cdot 1.96$				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) $\cdot 100$				
Analyst	NOTE: Expected weight = $\text{CVA} \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) \cdot 24) + (\text{TOC} - \text{ST}) / 100$				
12/07/98					
Analysis Date					
12/04/98	Sr-89/90 CONCENTRATION		1.63E-02	$\mu\text{Ci/mL}$	DETECTION LEVEL 1.71E-03 $\mu\text{Ci/mL}$
Analysis Time					
08:40 AM	RELATIVE COUNTING ERROR		16.4%		
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY		94.1%		

Analyst:	SCL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	7 Dec 98

SAMPLE.WB1 REV 2.0

22010NML

4677

LABCORE Completed Worklist Report for Worklist# 26968

Analyst: glt Instrument: AB10 Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: BY105 FOR @SR90-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	QSR90-01 SR90-01 SOLID		1.04E-03	1.02E-3	98.077 % Recovery	
1 STD	0	QSR90-01 SR90-01C SOLID		100	8.72E+01	87.200 % Recovery	
1 STD	0	QSR90-01 SR90-01E SOLID		1.00	2.11E+00	2.110 % Ct. Error	
2 BLNK-PREP	0	QSR90-01 SR90-01 SOLID		1	<1.48E-3	uCi/g	
2 BLNK-PREP	0	QSR90-01 SR90-01C SOLID		100	9.07E+01	90.700 % Recovery	
2 BLNK-PREP	0	QSR90-01 SR90-01E SOLID		1.00	3.43E+02	343.000 uCi/g	
3 BLNK/BKG	0	QSR90-01 SR90-01 SOLID		1	1.09E+00	1.090 BLNK/BKG	
4 SAMPLE	S98T002598	0 F QSR90-01 SR90-01 SOLID		N/A	1.44E-01	2.07E-003 uCi/g	
4 SAMPLE	S98T002598	0 F QSR90-01 SR90-01C SOLID		N/A	9.05E+01	% Recovery	
4 SAMPLE	S98T002598	0 F QSR90-01 SR90-01E SOLID		N/A	3.95E+00	% Ct. Error	
5 DUP	S98T002598	0 F QSR90-01 SR90-01 SOLID		1.44E-1	1.14E+0	155.140 RPD	
5 DUP	S98T002598	0 F QSR90-01 SR90-01C SOLID		100	9.07E+01	90.700 % Recovery	
5 DUP	S98T002598	0 F QSR90-01 SR90-01E SOLID		1.00	1.36E+00	1.360 % Ct. Err	
6 SAMPLE	S98T002599	0 F QSR90-01 SR90-01 SOLID		N/A	2.74E-01	2.07E-003 uCi/g	
6 SAMPLE	S98T002599	0 F QSR90-01 SR90-01C SOLID		N/A	8.90E+01	% Recovery	
6 SAMPLE	S98T002599	0 F QSR90-01 SR90-01E SOLID		N/A	2.82E+00	% Ct. Error	
7 DUP	S98T002599	0 F QSR90-01 SR90-01 SOLID		2.74E-1	2.05E-1	28.810 RPD	
7 DUP	S98T002599	0 F QSR90-01 SR90-01C SOLID		100	1.27E+02	127.000 % Recovery	
7 DUP	S98T002599	0 F QSR90-01 SR90-01E SOLID		1.00	2.73E+00	2.730 % Ct. Err	
8 SAMPLE	S98T002600	0 F QSR90-01 SR90-01 SOLID		N/A	6.91E+00	6.19E-003 uCi/g	
8 SAMPLE	S98T002600	0 F QSR90-01 SR90-01C SOLID		N/A	1.21E+02	% Recovery	
8 SAMPLE	S98T002600	0 F QSR90-01 SR90-01E SOLID		N/A	9.56E-01	% Ct. Error	
9 DUP	S98T002600	0 F QSR90-01 SR90-01 SOLID		6.91E+0	1.10E+1	45.673 RPD	
9 DUP	S98T002600	0 F QSR90-01 SR90-01C SOLID		100	8.74E+01	87.400 % Recovery	
9 DUP	S98T002600	0 F QSR90-01 SR90-01E SOLID		1.00	8.96E-01	0.896 % Ct. Err	

Final page for worklist# 26968

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____ Date _____

Samples will be held due to high RPDs. on 3/4/99
S98T002598 data included for informational purposes only.
DATA NOT USED

LABCORE Data Entry Template for Worklist# 26968

Analyst: gll Instrument: AB00 #10 Book# 28657Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105 FOR @SR90-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@SR90-01	SOLID		
2	BLNK-PREP				@SR90-01	SOLID		
3	BLNK/BKG				@SR90-01	SOLID		
4	SAMPLE	S98T002598	0	F	@SR90-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
5	DUP	S98T002598	0	F	@SR90-01	SOLID		
6	SAMPLE	S98T002599	0	F	@SR90-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
7	DUP	S98T002599	0	F	@SR90-01	SOLID		
8	SAMPLE	S98T002600	0	F	@SR90-01	SOLID	98000436	BY-105 (2)
		Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
9	DUP	S98T002600	0	F	@SR90-01	SOLID		

Final page for worklist # 26968

Analyst Signature [Signature] Date 11-8-98Analyst Signature [Signature] Date 11-9-98
Raymond 11-9-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

DATA
NOT USED

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	10	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	8911	GROSS WEIGHT (W2)	7.8815
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.7943
26968	BACKGROUND in cpm (BKG)	9.4	NET WEIGHT (W3)	0.0872
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	8.70
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR DDF	1		
LIQUID	SAMPLE COUNT RATE (Rs)	881.70	SR-90 EFFICIENCY FACTO (C1)	0.4051
Batch Number	CRITICAL LEVEL (Lc)	1.85	Y-90 EFFICIENCY FACTOR (C2)	0.4503
98005576	TIME OF SEPARATION (ST)	14:20	Rmax	N/A
Rerun	DATE OF SEPARATION (SD)	11/08/98	DETECTION LIMIT (Ld)	3.79
0	TIME OF COUNT (TOC)	23:02	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0224E+00
Sample Prep	DATE OF COUNT (DOC)	11/08/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0434E-03		
WL26968-STD				
Instrument Code				
WB26870	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF RS<=Lc AND RS>=0 OR REPLACE RS WITH Lc IF RS<0			
SLH2	RS*DF*DDF*1000/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))			
SAC	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT)*1.96)			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
GLL	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
11/09/98	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100			
Analysis Date				DETECTION LEVEL 4.39E-06 $\mu\text{Ci/L}$
11/08/98	Sr-89/90 CONCENTRATION	1.02E-03	$\mu\text{Ci/mL}$	
Analysis Time				
04:00 PM	RELATIVE COUNTING ERROR	2.1%		
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY	87.2%		

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9 Nov 98

STANDARD.WB1 REV 2.0

22010NML

DATA
NOT USED

4680

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	102	GROSS WEIGHT	(W2)	7.1562
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0655
26968	BACKGROUND in cpm	(BKG)	9.4	NET WEIGHT	(W3)	0.0907
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	8.88
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0372			
SOLID	SAMPLE COUNT RATE	(Rs)	0.80	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.85	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
98005576	TIME OF SEPARATION	(ST)	14:20	Rmax		2.71
Refun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	3.79
0	TIME OF COUNT	(TOC)	23:13	Sr-89/90 CONC in $\mu\text{Ci/g}$	<	1.4778E-03
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
N/A						
Sample #						

WL26968-BLNK						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} \geq 0$ or Replace RS with Lc if $\text{RS} < 0$					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						
11/08/98	Sr-89/90 CONCENTRATION		< 1.48E-03		$\mu\text{Ci/g}$	DETECTION LEVEL 2.07E-03 $\mu\text{Ci/g}$
Analysis Time	LESS THAN Value was Determined from Rmax.					
04:00 PM	RELATIVE COUNTING ERROR		343.0%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		90.7%			

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9-2-98

BLANK.WB1 REV 2.0

22010NML

DATA
NOT USED

4681

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)						SAMPLE	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000	
SAMPLE	TOTAL COUNTS	(TC)	2734	GROSS WEIGHT	(W2)	7.1625	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0720	
26968	BACKGROUND in cpm	(BKG)	9.4	NET WEIGHT	(W3)	0.0905	
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.05	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0372				
SOLID	SAMPLE COUNT RATE	(Rs)	264.00	SR-90 EFFICIENCY FACTOR	(C1)	0.4051	
Batch Number	CRITICAL LEVEL	(Lc)	1.85	Y-90 EFFICIENCY FACTOR	(C2)	0.4503	
98005576	TIME OF SEPARATION	(ST)	14:20	Rmax		N/A	
Rerun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	3.79	
0	TIME OF COUNT	(TOC)	23:23	Sr-89/90 CONC in µCi/g		1.4429E-01	
Sample Prep	DATE OF COUNT	(DOC)	11/08/98				
FUSION01							
Sample #							
S98T002598							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = ((The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/09/98							
Analysis Date							
11/08/98	Sr-89/90 CONCENTRATION		1.44E-01		µCi/g	DETECTION LEVEL 2.07E-03 µCi/g	
Analysis Time							
04:00 PM	RELATIVE COUNTING ERROR		3.9%				
Sample Point							
BY-105 (2)	PERCENT CARRIER RECOVERY		90.5%				

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9/14/98
SAMPLE.VB1 REV 2.0 22010NML			

DATA
NOT USED

4682

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	21034	GROSS WEIGHT	(W2)	7.1361
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0454
26968	BACKGROUND in cpm	(BKG)	9.4	NET WEIGHT	(W3)	0.0907
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.22
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0364	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
SOLID	SAMPLE COUNT RATE	(Rs)	2094.00	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
Batch Number	CRITICAL LEVEL	(Lc)	1.85	Rmax		N/A
98005576	TIME OF SEPARATION	(ST)	14:20	DETECTION LIMIT	(Ld)	3.79
Rerun	DATE OF SEPARATION	(SD)	11/08/98	Sr-89/90 CONC in µCi/g		1.1406E+00
0	TIME OF COUNT	(TOC)	23:33			
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
FUSION01						
Sample #						
S98T002598						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						
11/08/98	Sr-89/90 CONCENTRATION	1.14E+00		µCi/g	DETECTION LEVEL	
Analysis Time						
04:00 PM	RELATIVE COUNTING ERROR	1.4%		2.06E-03		
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	90.7%		µCi/g		

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9 Nov 98

SAMPLE.WB1 REV 2.0

22010NML

DATA
NOT USED

4683

LABCORE Completed Worklist Report for Worklist# 26969

Analyst: gl Instrument: AB12 Book# _____

Method: _____ Rev/Mod _____

Worklist Comment: BY105 FOR @SR90-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 SOLID		1.04E-03	1.06E-3	101.923 % Recovery	
1 STD	0	@SR90-01 SR90-01C SOLID		100	8.76E+01	87.600 % Recovery	
1 STD	0	@SR90-01 SR90-01E SOLID		1.00	2.07E+00	2.070 % Ct. Error	
2 BLNK-PREP	0	@SR90-01 SR90-01 SOLID		1	5.39E-2	0.054 uCi/g	
2 BLNK-PREP	0	@SR90-01 SR90-01C SOLID		100	8.80E+01	88.000 % Recovery	
2 BLNK-PREP	0	@SR90-01 SR90-01E SOLID		1.00	6.39E+00	6.390 uCi/g	
3 BLNK/BKG	0	@SR90-01 SR90-01 SOLID		1	8.14E+01	81.400 BLNK/BKG	
4 SAMPLE	S98T002601	0 F @SR90-01 SR90-01 SOLID		N/A	1.50E+00	7.77E-004 uCi/g	
4 SAMPLE	S98T002601	0 F @SR90-01 SR90-01C SOLID		N/A	8.89E+01	% Recovery	
4 SAMPLE	S98T002601	0 F @SR90-01 SR90-01E SOLID		N/A	1.19E+00	% Ct. Error	
5 DUP	S98T002601	0 F @SR90-01 SR90-01 SOLID		1.50E+0	1.47E+0	2.020 RPD	
5 DUP	S98T002601	0 F @SR90-01 SR90-01C SOLID		100	8.84E+01	88.400 % Recovery	
5 DUP	S98T002601	0 F @SR90-01 SR90-01E SOLID		1.00	1.20E+00	1.200 % Ct. Error	
6 SAMPLE	S98T002602	0 F @SR90-01 SR90-01 SOLID		N/A	6.20E+02	0.156 uCi/g	
6 SAMPLE	S98T002602	0 F @SR90-01 SR90-01C SOLID		N/A	8.80E+01	% Recovery	
6 SAMPLE	S98T002602	0 F @SR90-01 SR90-01E SOLID		N/A	8.28E-01	% Ct. Error	
7 DUP	S98T002602	0 F @SR90-01 SR90-01 SOLID		6.20E+2	6.38E+2	2.862 RPD	
7 DUP	S98T002602	0 F @SR90-01 SR90-01C SOLID		100	8.76E+01	87.600 % Recovery	
7 DUP	S98T002602	0 F @SR90-01 SR90-01E SOLID		1.00	8.11E-01	0.811 % Ct. Error	
8 SAMPLE	S98T002603	0 F @SR90-01 SR90-01 SOLID		N/A	7.69E+02	0.163 uCi/g	
8 SAMPLE	S98T002603	0 F @SR90-01 SR90-01C SOLID		N/A	8.63E+01	% Recovery	
8 SAMPLE	S98T002603	0 F @SR90-01 SR90-01E SOLID		N/A	7.61E-01	% Ct. Error	
9 DUP	S98T002603	0 F @SR90-01 SR90-01 SOLID		7.69E+2	8.10E+2	5.193 RPD	
9 DUP	S98T002603	0 F @SR90-01 SR90-01C SOLID		100	8.74E+01	87.400 % Recovery	
9 DUP	S98T002603	0 F @SR90-01 SR90-01E SOLID		1.00	7.42E-01	0.742 % Ct. Error	

Final page for worklist# 26969

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

[Signature]
Reviewer Signature Date

The blank appears contaminated, however, the activity in the blank is insignificant (3.6% of lowest activity sample) when compared to the sample. etc

Units shown for QC (BLK/BKG) may not reflect the actual units.

4684

11/04/98 15:41
x-0004-1

Page: 1

LABCORE Data Entry Template for Worklist# 26969

Analyst: gl Instrument: AB00 12 Book# 28357Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105 FOR @SR90-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002601 0 F		@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002601 0 F		@SR90-01	SOLID		
6 SAMPLE	S98T002602 0 F		@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
7 DUP	S98T002602 0 F		@SR90-01	SOLID		
8 SAMPLE	S98T002603 0 F		@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
9 DUP	S98T002603 0 F		@SR90-01	SOLID		

Final page for worklist # 26969

Day L. H. T. 11-8-98
 Analyst Signature Date

SL Hall 11-9-98
 Analyst Signature Date
W. Raymond 11-9-98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

Sr-89/90 : LA-220-01 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS					STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000	
STD	TOTAL COUNTS (TC)	8991	GROSS WEIGHT (W2)	7.1481	
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.0605	
26969	BACKGROUND in cpm (BKG)	1.2	NET WEIGHT (W3)	0.0876	
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	3.62	
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST DILUTION FACTOR DDF	1			
LIQUID	SAMPLE COUNT RATE (Rs)	897.90	SR-90 EFFICIENCY FACTO (C1)	0.4180	
Batch Number	CRITICAL LEVEL (Lc)	0.66	Y-90 EFFICIENCY FACTOR (C2)	0.4660	
98005577	TIME OF SEPARATION (ST)	10:35	Rmax	N/A	
Rerun	DATE OF SEPARATION (SD)	11/08/98	DETECTION LIMIT (Ld)	1.41	
0	TIME OF COUNT (TOC)	14:12	Sr-89/90 CONC. in µCi/L	1.0593E+00	
Sample Prep	DATE OF COUNT (DOC)	11/08/98			
N/A	STANDARD BOOK #	28B57			
Sample #	STANDARD VALUE in µCi/mL	1.0434E-03			
WL26969-STD					
Instrument Code					
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
Prepared By	Sr-89/90 CONC in µCi/mL REPLACE RS WITH RMAX IF RS<=Lc AND RS>=0 OR REPLACE RS WITH Lc IF RS<0				
SLH2	RS*DF*DDF*1000/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*REC*2220000)				
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
SAC	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT)*1.96)				
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
GLL	NOTE: Expected weight = CVA * 0.1				
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
11/09/98	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
Analysis Date				DETECTION LEVEL 1.66E-06 µCi/L	
11/08/98	Sr-89/90 CONCENTRATION	1.06E-03	µCi/mL		
Analysis Time					
01:10 PM	RELATIVE COUNTING ERROR	2.1%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	87.6%			

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9 Nov 98

STANDARD.WB1 REV 2.0

22010NML

4686

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	977	GROSS WEIGHT	(W2)	7.1632
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0752
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0880
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	5.37
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.064			
SOLID	SAMPLE COUNT RATE	(Rs)	96.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005577	TIME OF SEPARATION	(ST)	10:35	Rmax		N/A
Return	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41
0	TIME OF COUNT	(TOC)	15:57	Sr-89/90 CONC in µCi/g		5.3873E-02
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
N/A						
Sample #						

WL26969-BLNK

Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			DETECTION LEVEL 7.88E-04 µCi/g
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0			
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)			
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))			
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96			
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
Analyst	NOTE: Expected weight = CVA * 0.1			
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100			
11/09/98				
Analysis Date				
11/08/98	Sr-89/90 CONCENTRATION	5.39E-02	µCi/g	
Analysis Time				
01:10 PM	RELATIVE COUNTING ERROR	6.4%		
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY	88.0%		

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9 Nov 98

BLANK.WB1 REV 2.0

22010NML

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	27262	GROSS WEIGHT	(W2)	7.1685
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0796
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0889
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	5.67
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.084			
SOLID	SAMPLE COUNT RATE	(Rs)	2725.00	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4680
98005577	TIME OF SEPARATION	(ST)	10:35	Rmax		N/A
ReRun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41
0	TIME OF COUNT	(TOC)	16:15	Sr-89/90 CONC in $\mu\text{Ci/g}$		1.5011E+00
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
FUSION01						
Sample #						
S98T002601						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						DETECTION LEVEL 7.77E-04 $\mu\text{Ci/g}$
11/08/98	Sr-89/90 CONCENTRATION	1.50E+00	$\mu\text{Ci/g}$			
Analysis Time						
01:10 PM	RELATIVE COUNTING ERROR	1.2%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	88.9%				

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: <i>[Signature]</i>

SAMPLE.WB1 REV 2.0

22010NML

4688

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	26518	GROSS WEIGHT	(W2)	7.8720
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7836
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0884
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	5.98
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.06			
SOLID	SAMPLE COUNT RATE	(Rs)	2650.60	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005577	TIME OF SEPARATION	(ST)	10:35	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41
0	TIME OF COUNT	(TOC)	16:34	Sr-89/90 CONC in $\mu\text{Ci/g}$		1.4663E+00
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
FUSION01						
Sample #						
S98T002601						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*22200000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						DETECTION LEVEL 7.80E-04 $\mu\text{Ci/g}$
11/08/98	Sr-89/90 CONCENTRATION	1.47E+00	$\mu\text{Ci/g}$			
Analysis Time						
01:10 PM	RELATIVE COUNTING ERROR	1.2%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	88.4%				

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9 Nov 98

SAMPLE.WB1 REV 2.0

22010NML

4689

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	56024	GROSS WEIGHT	(W2) 7.1608
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.0728
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3) 0.0880
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT) 6.33
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0816		
SOLID	SAMPLE COUNT RATE	(Rs)	5601.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2) 0.4880
98005577	TIME OF SEPARATION	(ST)	10:40	Rmax	N/A
Retain	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld) 1.41
0	TIME OF COUNT	(TOC)	17:00	Sr-89/90 CONC in $\mu\text{Ci/g}$	6.1994E+02
Sample Prep	DATE OF COUNT	(DOC)	11/08/98		
FUSION01					
Sample #					
S98T002602					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/09/98					
Analysis Date					
11/08/98	Sr-89/90 CONCENTRATION	6.20E+02		$\mu\text{Ci/g}$	DETECTION LEVEL 1.56E-01 $\mu\text{Ci/g}$
Analysis Time					
01:10 PM	RELATIVE COUNTING ERROR	0.8%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	88.0%			

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:		SAC	Date: 9/14/98

SAMPLE.WB1 REV 2.0

22010NML

4690

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Sr-89/90 : LA-220-01 (E-3), 102 (INACTIVE), 104 (E-5)						DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	58469	GROSS WEIGHT	(W2)	7.1736
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0860
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0876
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	7.52
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.096			
SOLID	SAMPLE COUNT RATE	(Rs)	5845.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005577	TIME OF SEPARATION	(ST)	10:40	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41
0	TIME OF COUNT	(TOC)	18:11	Sr-89/90 CONC in $\mu\text{Ci/g}$		6.3764E+02
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
FUSION01						
Sample #						
S98T002602						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*D)))**SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						DETECTION LEVEL 1.54E-01 $\mu\text{Ci/g}$
11/08/98	Sr-89/90 CONCENTRATION		6.38E+02	$\mu\text{Ci/g}$		
Analysis Time						
01:10 PM	RELATIVE COUNTING ERROR		0.8%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		87.6%			

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	<i>[Signature]</i> SAC	Date:	<i>9 Nov 98</i>
SAMPLE.WB1 REV 2.0		22010NML	

4691

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	SAMPLE
SAMPLE	TOTAL COUNTS	(TC)	66356	GROSS WEIGHT	(W2)	7.8862
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.8019
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0863
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	7.78
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.998			
SOLID	SAMPLE COUNT RATE	(Rs)	6634.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005577	TIME OF SEPARATION	(ST)	10:40	Rmax		N/A
Run	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41
0	TIME OF COUNT	(TOC)	18:27	Sr-89/90 CONC in $\mu\text{Ci/g}$		7.6851E+02
Sample Prep	DATE OF COUNT	(DOC)	11/08/98			
FUSION01						
Sample #						
S98T002603						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/09/98						
Analysis Date						DETECTION
11/08/98	Sr-89/90 CONCENTRATION		7.69E+02		$\mu\text{Ci/g}$	LEVEL
Analysis Time						
01:10 PM	RELATIVE COUNTING ERROR		0.8%			1.63E-01
Sample Point						$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY		86.3%			

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	SAC	Date:	9/11/98

SAMPLE.WB1 REV 2.0

22010NML

4692

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000	
DUP	TOTAL COUNTS	(TC)	69834	GROSS WEIGHT	(W2)	7.1246	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0372	
26969	BACKGROUND in cpm	(BKG)	1.2	NET WEIGHT	(W3)	0.0874	
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	7.98	
@SR90-01	DILUTION FACTOR	DF	101				
Maldx	DIGEST FACTOR (g/L)	(D g/L)	1.9648				
SOLID	SAMPLE COUNT RATE	(Rs)	6982.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.66	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005577	TIME OF SEPARATION	(ST)	10:40	Rmax		N/A	
Refun	DATE OF SEPARATION	(SD)	11/08/98	DETECTION LIMIT	(Ld)	1.41	
0	TIME OF COUNT	(TOC)	18:39	Sr-89/90 CONC in $\mu\text{Ci/g}$		8.1047E+02	
Sample Prep	DATE OF COUNT	(DOC)	11/08/98				
FUSION01							
Sample #							
S98T002603							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/09/98							
Analysis Date							
11/08/98	Sr-89/90 CONCENTRATION					8.10E+02	DETECTION LEVEL 1.64E-01 $\mu\text{Ci/g}$
Analysis Time	RELATIVE COUNTING ERROR					0.7%	
01:10 PM	PERCENT CARRIER RECOVERY					87.4%	
Sample Point	BY-105 (2)						

Analyst:	GLL	Date:	09-Nov-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: <i>[Signature]</i>
SAMPLE.WB1 REV 2.0		22010NML	

LABCORE Completed Worklist Report for Worklist# 26970

Analyst: scl

Instrument: AB10

Book#

Method: Rev/Mod

Worklist Comment: BY105 FOR @SR90-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 SOLID		1.04E-03	1.12E-3	107.592 % Recovery	
1 STD	0	@SR90-01 SR90-01C SOLID		100	8.52E+01	85.200 % Recovery	
1 STD	0	@SR90-01 SR90-01E SOLID		1.00	2.02E+00	2.020 % Ct. Error	
2 BLNK-PREP	0	@SR90-01 SR90-01 SOLID		1	<1.10E-1	uCi/g	
2 BLNK-PREP	0	@SR90-01 SR90-01C SOLID		100	8.67E+01	86.700 % Recovery	
2 BLNK-PREP	0	@SR90-01 SR90-01E SOLID		1.00	5.00E+02	500.000 uCi/g	
3 BLNK/BKG	0	@SR90-01 SR90-01 SOLID		1	1.01E+00	1.010 BLNK/BKG	
4 SAMPLE	S98T002845	0 F @SR90-01 SR90-01 SOLID		N/A	2.80E+02	0.213 uCi/g	
4 SAMPLE	S98T002845	0 F @SR90-01 SR90-01C SOLID		N/A	8.59E+01	% Recovery	
4 SAMPLE	S98T002845	0 F @SR90-01 SR90-01E SOLID		N/A	8.88E-01	% Ct. Error	
5 DUP	S98T002845	0 F @SR90-01 SR90-01 SOLID		2.80E+2	2.93E+2	4.538 RPD	
5 DUP	S98T002845	0 F @SR90-01 SR90-01C SOLID		100	8.65E+01	86.500 % Recovery	
5 DUP	S98T002845	0 F @SR90-01 SR90-01E SOLID		1.00	8.69E-01	0.863 % Ct. Err	
6 SAMPLE	S98T002918	0 F @SR90-01 SR90-01 SOLID		N/A	1.95E+00	8.50E-003 uCi/g	
6 SAMPLE	S98T002918	0 F @SR90-01 SR90-01C SOLID		N/A	8.62E+01	% Recovery	
6 SAMPLE	S98T002918	0 F @SR90-01 SR90-01E SOLID		N/A	2.14E+00	% Ct. Error	
7 DUP	S98T002918	0 F @SR90-01 SR90-01 SOLID		1.95E+0	1.61E+0	19.101 RPD	
7 DUP	S98T002918	0 F @SR90-01 SR90-01C SOLID		100	8.20E+01	82.000 % Recovery	
7 DUP	S98T002918	0 F @SR90-01 SR90-01E SOLID		1.00	2.39E+00	2.390 % Ct. Err	
8 SAMPLE	S98T002919	0 F @SR90-01 SR90-01 SOLID		N/A	6.48E-01	2.06E-003 uCi/g	
8 SAMPLE	S98T002919	0 F @SR90-01 SR90-01C SOLID		N/A	8.68E+01	% Recovery	
8 SAMPLE	S98T002919	0 F @SR90-01 SR90-01E SOLID		N/A	1.82E+00	% Ct. Error	
9 DUP	S98T002919	0 F @SR90-01 SR90-01 SOLID		6.48E-1	9.17E-1	34.377 RPD	
9 DUP	S98T002919	0 F @SR90-01 SR90-01C SOLID		100	8.61E+01	86.100 % Recovery	
9 DUP	S98T002919	0 F @SR90-01 SR90-01E SOLID		1.00	1.54E+00	1.540 % Ct. Err	

Final page for worklist# 26970

Analyst Signature

Date

Analyst Signature

Date

Reviewer Signature

Date

Sample S98T002919 will be rerun due to RPD. Sample S98T002918 has high RPD (in limits), GEA shows Cs-137 RPD of 4850 ± ~19%, indicating possible sample inhomogeneity.

Units shown for QC (BLK/BKG) may not reflect the actual units.

Sample S98T002919 reanalyzed on WL# 27191. 3/4/99

LABCORE Data Entry Template for Worklist# 26970

Analyst: S.L. Instrument: AB00 10 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105 FOR @SR90-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002845	0 F	@SR90-01	SOLID	98000437	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002845	0 F	@SR90-01	SOLID		
6 SAMPLE	S98T002918	0 F	@SR90-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
7 DUP	S98T002918	0 F	@SR90-01	SOLID		
8 SAMPLE	S98T002919	0 F	@SR90-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
9 DUP	S98T002919	0 F	@SR90-01	SOLID		

Final page for worklist # 26970

Sue Lai 11-10-98
Analyst Signature Date

Sue Z. Cochran 11/11/98
Analyst Signature Date
MS 11/11/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

WORKBOOK PAGE: 17				STANDARD		
Sr-89/90 : LA-220-101 (E-3), 202 (INACTIVE), 104 (E-5) LIQUIDS						
Type	DETECTOR NUMBER	10	CARRIER ADDED in mL (CVA)	1.000		
STD	TOTAL COUNTS (TC)	9645	GROSS WEIGHT (W2)	7.8758		
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.7906		
26970	BACKGROUND in cpm (BKG)	9.1	NET WEIGHT (W3)	0.0852		
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	9.87		
@SR90-01	DILUTION FACTOR (DF)	1				
Matrix	DIGEST DILUTION FACTOR (DDF)	1				
LIQUID	SAMPLE COUNT RATE (Rs)	955.40	SR-90 EFFICIENCY FACTO (C1)	0.4051		
Batch Number	CRITICAL LEVEL (Lc)	1.82	Y-90 EFFICIENCY FACTOR (C2)	0.4503		
98005578	TIME OF SEPARATION (ST)	10:55	Rmax	N/A		
ReRun	DATE OF SEPARATION (SD)	11/10/98	DETECTION LIMIT (Ld)	3.73		
0	TIME OF COUNT (TOC)	20:47	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.1210E+00		
Sample Prep	DATE OF COUNT (DOC)	11/10/98				
N/A	STANDARD BOOK #	28B57				
Sample	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0433E-03				
WL26970						
Instrument Code						
WB26870	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS \geq 0$ OR REPLACE RS WITH Lc IF $RS < 0$					
SZC	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e^{-((-\text{natural log } 2) / 64.2 * DT))}) * SS * REC * 2220000)$					
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$					
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$					
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
SCL	NOTE: Expected weight = CVA * 0.1					
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
11/11/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$					
Analysis Date				DETECTION LEVEL 4.37E-06 $\mu\text{Ci/L}$		
11/10/98	Sr-89/90 CONCENTRATION	1.12E-03	$\mu\text{Ci/mL}$			
Analysis Time						
02:30 PM	RELATIVE COUNTING ERROR	2.0%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	85.2%				

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

STANDARD.WB1 REV 2.0

22010NML

4696

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	92	GROSS WEIGHT	(W2)	7.8636
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7769
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3)	0.0867
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	10.25
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.038			
SOLID	SAMPLE COUNT RATE	(Rs)	0.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
98005578	TIME OF SEPARATION	(ST)	10:55	Rmax		1.92
ReRun	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld)	3.73
0	TIME OF COUNT	(TOC)	21:10	Sr-89/90 CONC in $\mu\text{Ci/g}$	<	1.0959E-01
Sample Prep	DATE OF COUNT	(DOC)	11/10/98			
N/A						
Sample #						
WL26970						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS < Lc and RS >= 0 or Replace RS with Lc if RS < 0					
Prepared By	RS * 1000 * DF / ((C1 + C2 * (1 - e to the power of ((-natural log 2) / (64.2 * DT)))) * SS * (Dg/L) * REC * 2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/11/98						
Analysis Date						
11/10/98	Sr-89/90 CONCENTRATION		< 1.10E-01	$\mu\text{Ci/g}$		
Analysis Time	LESS THAN Value was Determined from Rmax.					
02:30 PM	RELATIVE COUNTING ERROR		500.0%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		86.7%			

DETECTION

LEVEL

2.12E-01

 $\mu\text{Ci/g}$

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

BLANK.WB1 REV 2.0

22010NML

4697

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

WORKBOOK PAGE: SR-89/90 : LA-220-01 (E-3), 102 (INACTIVE), 104 (E-5)					SAMPLE	
type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	49044	GROSS WEIGHT	(W2)	7.1174
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.0315
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3)	0.0859
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	10.65
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.038			
SOLID	SAMPLE COUNT RATE	(Rs)	4895.30	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
98005578	TIME OF SEPARATION	(ST)	10:55	Rmax		N/A
Refun	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld)	3.73
0	TIME OF COUNT	(TOC)	21:34	Sr-89/90 CONC in µCi/g 2.8021E+02		
Sample Prep	DATE OF COUNT	(DOC)	11/10/98			
FUSION01						
Sample #						
S98T002845						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT)))*SS*(Dg/L)*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/11/98						
Analysis Date						
11/10/98	Sr-89/90 CONCENTRATION	2.80E+02		µCi/g		DETECTION LEVEL 2.13E-01 µCi/g
Analysis Time						
02:30 PM	RELATIVE COUNTING ERROR	0.9%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	85.9%				

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

4698

WORKBOOK PAGE: DUPS

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

						DUP
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	51196	GROSS WEIGHT	(W2)	7.8775
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7910
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3)	0.0865
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	10.98
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0152			
SOLID	SAMPLE COUNT RATE	(Rs)	5110.50	Sr-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
98005578	TIME OF SEPARATION	(ST)	10:55	Rmax		N/A
Regrin	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld)	3.73
0	TIME OF COUNT	(TOC)	21:54	Sr-89/90 CONC in µCi/g		2.9286E+02
Sample Prep	DATE OF COUNT	(DOC)	11/10/98			
FUSION01						
Sample #						
S98T002845						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.98					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/11/98						
Analysis Date						DETECTION LEVEL 2.13E-01 µCi/g
11/10/98	Sr-89/90 CONCENTRATION		2.93E+02	µCi/g		
Analysis Time						
02:30 PM	RELATIVE COUNTING ERROR		0.9%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		86.5%			

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

4699

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)						SAMPLE	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000	
SAMPLE	TOTAL COUNTS	(TC)	8639	GROSS WEIGHT	(W2)	7.8698	
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7836	
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3)	0.0862	
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	11.13	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0084				
SOLID	SAMPLE COUNT RATE	(Rs)	854.80	SR-90 EFFICIENCY FACTOR	(C1)	0.4051	
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2)	0.4503	
98005578	TIME OF SEPARATION	(ST)	11:00	Rmax		N/A	
Retent	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld)	3.73	
0	TIME OF COUNT	(TOC)	22:08	Sr-89/90 CONC in µCi/g	1.9505E+00		
Sample Prep	DATE OF COUNT	(DOC)	11/10/98				
FUSION01							
Sample#							
S98T002918							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/(((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.98						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/11/98							
Analysis Date							
11/10/98	Sr-89/90 CONCENTRATION	1.95E+00		µCi/g	DETECTION LEVEL 8.50E-03 µCi/g		
Analysis Time							
02:30 PM	RELATIVE COUNTING ERROR	2.1%					
Sample Point							
BY-105 (2)	PERCENT CARRIER RECOVERY	86.2%					

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

4700

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	6999	GROSS WEIGHT	(W2) 7.1333
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.0513
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3) 0.0820
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 12.35
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0384		
SOLID	SAMPLE COUNT RATE	(Rs)	690.80	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
98005578	TIME OF SEPARATION	(ST)	11:00	Rmax	N/A
Refun	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld) 3.73
0	TIME OF COUNT	(TOC)	23:21	Sr-89/90 CONC in µCi/g	1.6142E+00
Sample Prep	DATE OF COUNT	(DOC)	11/10/98		
FUSION01					
Sample #					
S98T002918					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT)))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.98				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/11/98					
Analysis Date					DETECTION LEVEL 8.71E-03 µCi/g
11/10/98	Sr-89/90 CONCENTRATION	1.61E+00	µCi/g		
Analysis Time					
02:30 PM	RELATIVE COUNTING ERROR	2.4%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	82.0%			

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	
SAMPLE:WB1 REV 2.0 22010NML			

4701

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	11820	GROSS WEIGHT	(W2)	7.8713
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7845
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3)	0.0868
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	14.53
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.996			
SOLID	SAMPLE COUNT RATE	(Rs)	1172.90	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
98005578	TIME OF SEPARATION	(ST)	11:00	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld)	3.73
0	TIME OF COUNT	(TOC)	01:32	Sr-89/90 CONC in $\mu\text{Ci/g}$ 6.4815E-01		
Sample Prep	DATE OF COUNT	(DOC)	11/11/98			
FUSION01						
Sample #						
S98T002919						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	$RS * 1000 * DF / ((C1 + C2 * (1 - e \text{ to the power of } ((-\text{natural log } 2) / 64.2 * DT))) * SS * (Dg/L) * REC * 2220000)$					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$					
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = $CVA * 0.1$					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$					
11/11/98						
Analysis Date						DETECTION LEVEL 2.06E-03 $\mu\text{Ci/g}$
11/10/98	Sr-89/90 CONCENTRATION	6.48E-01	$\mu\text{Ci/g}$			
Analysis Time						
02:30 PM	RELATIVE COUNTING ERROR	1.8%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	86.8%				

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

DATA
NOT USED

4702

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	16382	GROSS WEIGHT	(W2) 7.8773
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7912
26970	BACKGROUND in cpm	(BKG)	9.1	NET WEIGHT	(W3) 0.0861
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 14.87
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.97		
SOLID	SAMPLE COUNT RATE	(Rs)	1629.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.82	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
98005578	TIME OF SEPARATION	(ST)	11:00	Rmax	N/A
ReRun	DATE OF SEPARATION	(SD)	11/10/98	DETECTION LIMIT	(Ld) 3.73
0	TIME OF COUNT	(TOC)	01:52	Sr-89/90 CONC in µCi/g	9.1685E-01
Sample Prep	DATE OF COUNT	(DOC)	11/11/98		
FUSION01					
Sample #					
S98T002919					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26370	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/11/98					
Analysis Date					DETECTION LEVEL
11/10/98	Sr-89/90 CONCENTRATION	9.17E-01	µCi/g		
Analysis Time					
02:30 PM	RELATIVE COUNTING ERROR	1.5%		2.10E-03	
Sample Point					µCi/g
BY-105 (2)	PERCENT CARRIER RECOVERY	86.1%			

Analyst:	SCL	Date:	11-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

DATA
NOT USED

4703

LABCORE Completed Worklist Report for Worklist# 27091

Analyst: scl Instrument: AB10 Book#

Method: Rev/Mod

Worklist Comment: RERUN OF 26968,BY-105 szc

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		QSR90-01 SR90-01 SOLID		1.04E-03	1.12E-3	107.692	% Recovery
1 STD	0		QSR90-01 SR90-01C SOLID		100	8.43E+01	84.300	% Recovery
1 STD	0		QSR90-01 SR90-01E SOLID		1.00	2.03E+00	2.030	% Ct. Error
2 BLNK-PREP	0		QSR90-01 SR90-01 SOLID		1	1.95E-3	0.002	uCi/g
2 BLNK-PREP	0		QSR90-01 SR90-01C SOLID		100	8.25E+01	82.500	% Recovery
2 BLNK-PREP	0		QSR90-01 SR90-01E SOLID		1.00	8.79E+01	87.900	uCi/g
3 BLNK/BKG	0		QSR90-01 SR90-01 SOLID		1	1.35E+00	1.350	BLNK/BKG
4 SAMPLE	S98T002598	0 F	QSR90-01 SR90-01 SOLID		N/A	1.57E-01	2.17E-003	uCi/g
4 SAMPLE	S98T002598	0 F	QSR90-01 SR90-01C SOLID		N/A	8.45E+01		% Recovery
4 SAMPLE	S98T002598	0 F	QSR90-01 SR90-01E SOLID		N/A	3.88E+00		% Ct. Error
5 DUP	S98T002598	0 F	QSR90-01 SR90-01 SOLID		1.57E-1	1.14E+0	151.581	RPD
5 DUP	S98T002598	0 F	QSR90-01 SR90-01C SOLID		100	8.53E+01	85.300	% Recovery
5 DUP	S98T002598	0 F	QSR90-01 SR90-01E SOLID		1.00	1.39E+00	1.390	% Ct. Error
6 SAMPLE	S98T002599	0 F	QSR90-01 SR90-01 SOLID		N/A	2.75E-01	2.09E-003	uCi/g
6 SAMPLE	S98T002599	0 F	QSR90-01 SR90-01C SOLID		N/A	8.58E+01		% Recovery
6 SAMPLE	S98T002599	0 F	QSR90-01 SR90-01E SOLID		N/A	2.84E+00		% Ct. Error
7 DUP	S98T002599	0 F	QSR90-01 SR90-01 SOLID		2.75E-1	3.06E-1	10.671	RPD
7 DUP	S98T002599	0 F	QSR90-01 SR90-01C SOLID		100	8.57E+01	85.700	% Recovery
7 DUP	S98T002599	0 F	QSR90-01 SR90-01E SOLID		1.00	2.69E+00	2.690	% Ct. Error
8 SAMPLE	S98T002600	0 F	QSR90-01 SR90-01 SOLID		N/A	9.75E+00	8.67E-003	uCi/g
8 SAMPLE	S98T002600	0 F	QSR90-01 SR90-01C SOLID		N/A	8.44E+01		% Recovery
8 SAMPLE	S98T002600	0 F	QSR90-01 SR90-01E SOLID		N/A	9.54E-01		% Ct. Error
9 DUP	S98T002600	0 F	QSR90-01 SR90-01 SOLID		9.75E+0	1.09E+1	11.138	RPD
9 DUP	S98T002600	0 F	QSR90-01 SR90-01C SOLID		100	8.40E+01	84.000	% Recovery
9 DUP	S98T002600	0 F	QSR90-01 SR90-01E SOLID		1.00	9.00E-01	0.900	% Ct. Error

Final page for worklist# 27091

Analyst Signature Date Analyst Signature Date

[Signature] 12/7/98
Reviewer Signature Date

Results for S98T002598 are similar to the first run. Results are almost a factor of 10 different, most likely due to sample inhomogeneity. Results accepted, see.

Units shown for QC (BLK/BKG) may not reflect the actual units.

See WL# 26968 for results from first analysis on S98T002598. 3/14/99

LABCORE Data Entry Template for Worklist# 27091

Analyst: S.L. Instrument: AB00 10 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: RERUN OF 26968,BY-105 szc

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002598	0 F	@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002598	0 F	@SR90-01	SOLID		
6 SAMPLE	S98T002599	0 F	@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
7 DUP	S98T002599	0 F	@SR90-01	SOLID		
8 SAMPLE	S98T002600	0 F	@SR90-01	SOLID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
9 DUP	S98T002600	0 F	@SR90-01	SOLID		

Final page for worklist # 27091

Sue Li 11-12-98
Analyst Signature Date

Sharon L. Under 11/13/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

Sr-89/90 : LA-220-101 (E-3), 202 (INACTIVE), 104 (E-5) LIQUIDS				STANDARD		
Type	DETECTOR NUMBER	10	CARRIER ADDED in mL (CVA)	1.000		
STD	TOTAL COUNTS (TC)	9573	GROSS WEIGHT (W2)	7.3954		
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3111		
27091	BACKGROUND in cpm (BKG)	9.3	NET WEIGHT (W3)	0.0843		
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	10.00		
@SR90-01	DILUTION FACTOR (DF)	1				
Matrix	DIGEST DILUTION FACTOR DDF	1				
LIQUID	SAMPLE COUNT RATE (Rs)	948.00	SR-90 EFFICIENCY FACTO (C1)	0.4051		
Batch Number	CRITICAL LEVEL (Lc)	1.84	Y-90 EFFICIENCY FACTOR (C2)	0.4503		
98005699	TIME OF SEPARATION (ST)	11:08	Rmax	N/A		
ReRun	DATE OF SEPARATION (SD)	11/12/98	DETECTION LIMIT (Ld)	3.77		
0	TIME OF COUNT (TOC)	21:08	Sr-89/90 CONC. in $\mu\text{Ci/L}$ 1.1227E+00			
Sample Prep	DATE OF COUNT (DOC)	11/12/98				
N/A	STANDARD BOOK #	28B57				
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0431E-03				
WL27091						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG) Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $\text{RS} \leq \text{Lc}$ AND $\text{RS} \geq 0$ OR REPLACE RS WITH Lc If $\text{RS} < 0$ $\text{RS} * \text{DF} * \text{DDF} * 1000 / ((\text{C1} + \text{C2} * (1 - e \text{ to the power of } ((-\text{natural log } 2) / 64.2 * \text{DT})))) * \text{SS} * \text{REC} * 2220000$ NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$ Relative Counting Error = The Square Root of $((\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT}) * 1.96)$ Percent Carrier Recovery = (Net Weight / Expected weight) * 100 NOTE: Expected weight = $\text{CVA} * 0.1$ Detection Levels and Less Than Values are determined from Procedure LA-508-002. Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$					
WB26870						
Prepared By						
SZC						
Chemist						
SAC						
Analyst						
SCL						
Date Complete						
11/13/98						
Analysis Date						
11/12/98					Sr-89/90 CONCENTRATION	1.12E-03
Analysis Time						
02:45 PM	RELATIVE COUNTING ERROR	2.0%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	84.3%				

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12/Nov/98

STANDARD.WB1 REV 2.0

22010NNL

4706

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000	
BLNK-PREP	TOTAL COUNTS	(TC)	126	GROSS WEIGHT	(W2)	7.3915	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3090	
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3)	0.0825	
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	10.45	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0372				
SOLID	SAMPLE COUNT RATE	(Rs)	3.30	SR-90 EFFICIENCY FACTOR	(C1)	0.4051	
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2)	0.4503	
98005699	TIME OF SEPARATION	(ST)	11:08	Rmax		N/A	
Refun	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld)	3.77	
0	TIME OF COUNT	(TOC)	21:35	Sr-89/90 CONC in µCi/g		1.9518E-03	
Sample Prep	DATE OF COUNT	(DOC)	11/12/98				
N/A							
Sample #							
WL27091							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/13/98							
Analysis Date							
11/12/98	Sr-89/90 CONCENTRATION		1.95E-03	µCi/g	DETECTION LEVEL 2.23E-03 µCi/g		
Analysis Time							
02:45 PM	RELATIVE COUNTING ERROR		87.9%				
Sample Point							
BY-105 (2)	PERCENT CARRIER RECOVERY		82.5%				

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12 Nov 98

BLANK.WB1 REV 2.0

22010NML

4707

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	2819	GROSS WEIGHT	(W2) 7.4277
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3432
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3) 0.0845
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 10.65
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0372		
SOLID	SAMPLE COUNT RATE	(Rs)	272.60	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
98005699	TIME OF SEPARATION	(ST)	11:08	Rmax	N/A
Run#	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld) 3.77
1	TIME OF COUNT	(TOC)	21:47	Sr-89/90 CONC in $\mu\text{Ci/g}$	1.5711E-01
Sample Rep	DATE OF COUNT	(DOC)	11/12/98		
FUSION01					
Sample #					
S98T002598					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * 1000 * DF / ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / (64.2 * DT)))) * SS * (Dg/L) * REC * 2220000)$				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $CVA * 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
11/13/98					
Analysis Date					
11/12/98	Sr-89/90 CONCENTRATION	1.57E-01	$\mu\text{Ci/g}$	DETECTION LEVEL 2.17E-03 $\mu\text{Ci/g}$	
Analysis Time					
02:45 PM	RELATIVE COUNTING ERROR	3.9%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	84.5%			

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	17 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4708

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000	DUP
DUP	TOTAL COUNTS	(TC)	20150	GROSS WEIGHT	(W2)	7.4091	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3238	
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3)	0.0853	
Task Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	10.85	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0364				
SOLID	SAMPLE COUNT RATE	(Rs)	2005.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4051	
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2)	0.4503	
98005699	TIME OF SEPARATION	(ST)	11:08	Rmax		N/A	
ReRun	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld)	3.77	
1	TIME OF COUNT	(TOC)	21:59	Sr-89/90 CONC in µCi/g		1.1434E+00	
Sample Prep	DATE OF COUNT	(DOC)	11/12/98				
FUSION01							
Sample #							
S98T002598							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*(Dg/L)*REC*2220000)						
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/13/98							
Analysis Date							
11/12/98	Sr-89/90 CONCENTRATION		1.14E+00		µCi/g		DETECTION LEVEL 2.15E-03 µCi/g
Analysis Time							
02:45 PM	RELATIVE COUNTING ERROR		1.4%				
Sample Count							
BY-105 (2)	PERCENT CARRIER RECOVERY		85.3%				

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	11/12/98
SAMPLE.WB1 REV 2.0		22010NML	

4709

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	5034	GROSS WEIGHT	(W2) 7.3851
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.2993
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3) 0.0858
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 11.57
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0608		
SOLID	SAMPLE COUNT RATE	(Rs)	494.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
98005699	TIME OF SEPARATION	(ST)	11:13	Rmax	N/A
Run	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld) 3.77
1	TIME OF COUNT	(TOC)	22:47	Sr-89/90 CONC in $\mu\text{Ci/g}$	2.7486E-01
Sample Prep	DATE OF COUNT	(DOC)	11/12/98		
FUSION01					
Sample #					
S98T002599					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} * 1000 * \text{DF} / ((\text{C1} + \text{C2} * (1 - e \text{ to the power of } ((-\text{natural log } 2) / (64.2 * \text{DT})))) * \text{SS} * (\text{Dg/L}) * \text{REC} * 2220000)$				
SZC	NOTE: $64.2 = \text{Half Life for Y-90 and Rec.} = \text{Fractional Carrier Recovery } ((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$				
11/13/98					DETECTION LEVEL 2.09E-03 $\mu\text{Ci/g}$
Analysis Date					
11/12/98	Sr-89/90 CONCENTRATION	2.75E-01	$\mu\text{Ci/g}$		
Analysis Time					
02:45 PM	RELATIVE COUNTING ERROR	2.8%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	85.8%			

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4710

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	5594	GROSS WEIGHT	(W2) 7.3854
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.2997
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3) 0.0857
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 11.78
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0564		
SOLID	SAMPLE COUNT RATE	(Rs)	550.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
98005699	TIME OF SEPARATION	(ST)	11:13	Rmax	N/A
ReRun	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld) 3.77
1	TIME OF COUNT	(TOC)	23:00	Sr-89/90 CONC in µCi/g	3.0640E-01
Sample Prep	DATE OF COUNT	(DOC)	11/12/98		
FUSION01					
Sample #					
S98T002599					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/13/98					
Analysis Date					DETECTION LEVEL 2.10E-03 µCi/g
11/12/98	Sr-89/90 CONCENTRATION	3.06E-01	µCi/g		
Analysis Time					
02:45 PM	RELATIVE COUNTING ERROR	2.7%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	85.7%			

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4711

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000	
SAMPLE	TOTAL COUNTS	(TC) 42447		GROSS WEIGHT	(W2) 7.3831	
Work List	COUNT TIME in MINUTES	(CT) 10		TARE WEIGHT	(W1) 7.2987	
27091	BACKGROUND in cpm	(BKG) 9.3		NET WEIGHT	(W3) 0.0844	
Test Code	SAMPLE VOLUME in mL	(SS) 0.250		DELTA TIME (HOURS)	(DT) 12.02	
@SR90-01	DILUTION FACTOR	DF 1				
Matrix	DIGEST FACTOR (g/L)	(D g/L) 2.0164				
SOLID	SAMPLE COUNT RATE	(Rs) 4235.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4051	
Batch Number	CRITICAL LEVEL	(Lc) 1.84	Y-90 EFFICIENCY FACTOR	(C2)	0.4503	
98005699	TIME OF SEPARATION	(ST) 11:13	Rmax		N/A	
Rerun	DATE OF SEPARATION	(SD) 11/12/98	DETECTION LIMIT	(Ld)	3.77	
1	TIME OF COUNT	(TOC) 23:14	Sr-89/90 CONC in $\mu\text{Ci/g}$		9.7505E+00	
Sample Prep	DATE OF COUNT	(DOC) 11/12/98				
FUSION01						
Sample #						
S98T002600						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	$RS * 1000 * DF / ((C1 + C2) * (1 - e \text{ to the power of } ((- \text{natural log } 2) / 64.2 * DT))) * SS * (Dg/L) * REC * 2220000$					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$					
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$					
11/13/98						
Analysis Date						
11/12/98	Sr-89/90 CONCENTRATION	9.75E+00	$\mu\text{Ci/g}$	DETECTION LEVEL 8.67E-03 $\mu\text{Ci/g}$		
Analysis Time						
02:45 PM	RELATIVE COUNTING ERROR	1.0%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	84.4%				

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12 Nov 98

SAMPLE.WB1 REV 2.0

22010NML

4712

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000	
DUP	TOTAL COUNTS	(TC)	47706	GROSS WEIGHT	(W2) 7.4372	
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3532	
27091	BACKGROUND in cpm	(BKG)	9.3	NET WEIGHT	(W3) 0.0840	
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 13.57	
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0084			
SOLID	SAMPLE COUNT RATE	(Rs)	4761.30	SR-90 EFFICIENCY FACTOR	(C1) 0.4051	
Batch Number	CRITICAL LEVEL	(Lc)	1.84	Y-90 EFFICIENCY FACTOR	(C2) 0.4503	
98005699	TIME OF SEPARATION	(ST)	11:13	Rmax	N/A	
ReRun	DATE OF SEPARATION	(SD)	11/12/98	DETECTION LIMIT	(Ld) 3.77	
1	TIME OF COUNT	(TOC)	00:47	Sr-89/90 CONC in µCi/g	1.0902E+01	
Sample Prep	DATE OF COUNT	(DOC)	11/13/98			
FUSION01						
Sample #						
S98T002600						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/13/98						
Analysis Date						
11/12/98	Sr-89/90 CONCENTRATION	1.09E+01		µCi/g	DETECTION LEVEL 8.62E-03 µCi/g	
Analysis Time						
02:45 PM	RELATIVE COUNTING ERROR	0.9%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	84.0%				

Analyst:	SCL	Date:	13-Nov-98
Signature of Chemist:	SAC	Date:	12 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4713

LABCORE Completed Worklist Report for Worklist# 27136

Analyst: gll Instrument: AB12 Book# _____
Method: _____ Rev/Mod _____
Worklist Comment: BY105, @SR90-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@SR90-01 SR90-01 SOLID		1.04E-03	1.01E-3	97.115 %	Recovery
1 STD	0		@SR90-01 SR90-01C SOLID		100	9.37E+01	93.700 %	Recovery
1 STD	0		@SR90-01 SR90-01E SOLID		1.00	2.05E+00	2.050 %	Ct. Err
2 BLNK-PREP	0		@SR90-01 SR90-01 SOLID		1	6.66E-2	0.067	uCi/g
2 BLNK-PREP	0		@SR90-01 SR90-01C SOLID		100	9.21E+01	92.100 %	Recovery
2 BLNK-PREP	0		@SR90-01 SR90-01E SOLID		1.00	2.18E+01	21.800	uCi/g
3 BLNK/BKG	0		@SR90-01 SR90-01 SOLID		1	5.50E+00	5.500	BLNK/BKG
4 SAMPLE	S98T002712	0 F	@SR90-01 SR90-01 SOLID		N/A	4.64E+01	1.16E-002	uCi/g
4 SAMPLE	S98T002712	0 F	@SR90-01 SR90-01C SOLID		N/A	9.18E+01		% Recovery
4 SAMPLE	S98T002712	0 F	@SR90-01 SR90-01E SOLID		N/A	6.87E-01		% Ct. Error
5 DUP	S98T002712	0 F	@SR90-01 SR90-01 SOLID		4.64E+1	4.55E+1	1.959	RPD
5 DUP	S98T002712	0 F	@SR90-01 SR90-01C SOLID		100	9.14E+01	91.400 %	Recovery
5 DUP	S98T002712	0 F	@SR90-01 SR90-01E SOLID		1.00	6.95E-01	0.695	% Ct. Err
6 SAMPLE	S98T002739	0 F	@SR90-01 SR90-01 SOLID		N/A	3.83E+00	2.27E-003	uCi/g
6 SAMPLE	S98T002739	0 F	@SR90-01 SR90-01C SOLID		N/A	9.16E+01		% Recovery
6 SAMPLE	S98T002739	0 F	@SR90-01 SR90-01E SOLID		N/A	1.06E+00		% Ct. Error
7 DUP	S98T002739	0 F	@SR90-01 SR90-01 SOLID		3.83E+0	3.96E+0	3.338	RPD
7 DUP	S98T002739	0 F	@SR90-01 SR90-01C SOLID		100	9.17E+01	91.700 %	Recovery
7 DUP	S98T002739	0 F	@SR90-01 SR90-01E SOLID		1.00	1.04E+00	1.040	% Ct. Err
8 SAMPLE	S98T002844	0 F	@SR90-01 SR90-01 SOLID		N/A	3.63E+01	1.14E-002	uCi/g
8 SAMPLE	S98T002844	0 F	@SR90-01 SR90-01C SOLID		N/A	9.21E+01		% Recovery
8 SAMPLE	S98T002844	0 F	@SR90-01 SR90-01E SOLID		N/A	7.70E-01		% Ct. Error
9 DUP	S98T002844	0 F	@SR90-01 SR90-01 SOLID		3.63E+1	3.66E+1	0.823	RPD
9 DUP	S98T002844	0 F	@SR90-01 SR90-01C SOLID		100	9.17E+01	91.700 %	Recovery
9 DUP	S98T002844	0 F	@SR90-01 SR90-01E SOLID		1.00	7.66E-01	0.766	% Ct. Err

Final page for worklist# 27136

Analyst Signature	Date	Analyst Signature	Date
Reviewer Signature	Date		
<i>SA Carlson</i>	8 Dec 98		

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27136

Analyst: gll Instrument: AB00 #12 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105, @SR90-01, tdm

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@SR90-01	SOLID		
2	BLNK-PREP				@SR90-01	SOLID		
3	BLNK/BKG				@SR90-01	SOLID		
4	SAMPLE	S98T002712	0	F	@SR90-01	SOLID	98000436	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E								
5	DUP	S98T002712	0	F	@SR90-01	SOLID		
6	SAMPLE	S98T002739	0	F	@SR90-01	SOLID	98000490	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E								
7	DUP	S98T002739	0	F	@SR90-01	SOLID		
8	SAMPLE	S98T002844	0	F	@SR90-01	SOLID	98000437	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E								
9	DUP	S98T002844	0	F	@SR90-01	SOLID		

Final page for worklist # 27136

Ray LeBlond 12-7-98
Analyst Signature Date

W. Raymond 12-7-98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD	
TYPE	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)		1.000
STD	TOTAL COUNTS (TC)	9209	GROSS WEIGHT (W2)		7.4352
Work list	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.3415
27136	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT (W3)		0.0937
Isot Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)		3.58
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	918.30	SR-90 EFFICIENCY FACTO (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98005753	TIME OF SEPARATION (ST)	09:50	Rmax		N/A
Return	DATE OF SEPARATION (SD)	12/07/98	DETECTION LIMIT (Ld)		2.03
0	TIME OF COUNT (TOC)	13:25	Sr-89/90 CONC. in $\mu\text{Ci/L}$		1.0133E+00
Sample Prep	DATE OF COUNT (DOC)	12/07/98			
N/A	STANDARD BOOK #	28B57			
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0414E-03			
STD					
Instrument Code					
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS > 0$ OR REPLACE RS WITH Lc IF $RS < 0$				
VAR	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e^{-((-\text{natural log } 2) / (64.2 * DT))})) * SS * REC * 2220000)$				
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$				
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
GLL	NOTE: Expected weight = CVA * 0.1				
Data Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
12/07/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
Analysis Date					DETECTION LEVEL 2.24E-06 $\mu\text{Ci/L}$
12/07/98	Sr-89/90 CONCENTRATION	1.01E-03	$\mu\text{Ci/mL}$		
Analysis time					
12:25 PM	RELATIVE COUNTING ERROR	2.1%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	93.7%			

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: <i>8 Dec 98</i>

STANDARD.WB1 REV 2.0

22010NML

4716

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	143	GROSS WEIGHT	(W2)	7.3876
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.2955
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	0.0921
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	3.77
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9668			
SOLID	SAMPLE COUNT RATE	(Rs)	11.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005753	TIME OF SEPARATION	(ST)	09:50	Rmax		N/A
Run	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT	(TOC)	13:36	Sr-89/90 CONC in µCi/g		6.6644E-02
Sample Prep	DATE OF COUNT	(DOC)	12/07/98			
N/A						
Sample						
BLNK						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						
12/07/98	Sr-89/90 CONCENTRATION	6.66E-02		µCi/g		DETECTION LEVEL
Analysis Time						
12:25 PM	RELATIVE COUNTING ERROR	21.8%				1.16E-02
Sample Point						µCi/g
BY-105 (2)	PERCENT CARRIER RECOVERY	92.1%				

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	<i>SA Carter</i>	SAC	Date: 8 Dec 98

BLANK.WB1 REV 2.0

22010NML

4717

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000	SAMPLE
SAMPLE	TOTAL COUNTS	(TC)	81441	GROSS WEIGHT	(W2)	7.4006	
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3088	
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	0.0918	
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	4.00	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9668				
SOLID	SAMPLE COUNT RATE	(Rs)	8141.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005753	TIME OF SEPARATION	(ST)	09:50	Rmax		N/A	
Remd	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld)	2.03	
0	TIME OF COUNT	(TOC)	13:50	Sr-89/90 CONC in $\mu\text{Ci/g}$		4.6406E+01	
Sample Prep	DATE OF COUNT	(DOC)	12/07/98				
FUSION01							
Sample #							
S98T002712							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS < Lc$ and $RS >= 0$ or Replace RS with Lc if $RS < 0$						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*22200000)						
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
12/07/98							
Analysis Date							
12/07/98	Sr-89/90 CONCENTRATION		4.64E+01		$\mu\text{Ci/g}$		DETECTION LEVEL
Analysis Time							
12:25 PM	RELATIVE COUNTING ERROR		0.7%				1.16E-02
Sample Point							$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY		91.8%				

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	8 Dec 98
SAMPLE.WB1 REV 2.0	22010NML		

4718

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	79528	GROSS WEIGHT	(W2) 7.3834
Worksheet	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.2920
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3) 0.0914
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 4.22
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9616		
SOLID	SAMPLE COUNT RATE	(Rs)	7950.00	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005753	TIME OF SEPARATION	(ST)	09:50	Rmax	N/A
Return	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld) 2.03
0	TIME OF COUNT	(TOC)	14:03	Sr-89/90 CONC in $\mu\text{Ci/g}$	4.5525E+01
Sample Prep	DATE OF COUNT	(DOC)	12/07/98		
FUSION01					
Sample #					
S98T002712					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL
12/07/98	4.55E+01				$\mu\text{Ci/g}$
Analysis Time					
12:25 PM	RELATIVE COUNTING ERROR				1.16E-02
Sample Point					$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY				91.4%

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	<i>SAC</i>	Date:	8 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4719

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000	
SAMPLE	TOTAL COUNTS (TC)	34299	GROSS WEIGHT	(W2)	7.4183	
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3267	
27136	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT	(W3)	0.0916	
Test Code	SAMPLE VOLUME in mL (SS)	0.500	DELTA TIME (HOURS)	(DT)	4.30	
@SR90-01	DILUTION FACTOR (DF)	1				
Matrix	DIGEST FACTOR (g/L) (D g/L)	2.0048				
SOLID	SAMPLE COUNT RATE (Rs)	3427.30	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005753	TIME OF SEPARATION (ST)	09:55	Rmax		N/A	
Refun	DATE OF SEPARATION (SD)	12/07/98	DETECTION LIMIT	(Ld)	2.03	
0	TIME OF COUNT (TOC)	14:13	Sr-89/90 CONC in µCi/g		3.8288E+00	
Sample Prep	DATE OF COUNT (DOC)	12/07/98				
FUSION01						
Sample #						
S98T002739						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*(Dg/L)*REC*2220000)					
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/07/98						
Analysis Date						
12/07/98	Sr-89/90 CONCENTRATION	3.83E+00	µCi/g	DETECTION LEVEL 2.27E-03 µCi/g		
Analysis Time						
12:25 PM	RELATIVE COUNTING ERROR	1.1%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	91.6%				

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	8 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4720

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	35674	GROSS WEIGHT	(W2) 7.3787
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.2870
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3) 0.0917
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT) 4.47
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0124		
SOLID	SAMPLE COUNT RATE	(Rs)	3564.80	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005753	TIME OF SEPARATION	(ST)	09:55	Rmax	N/A
Refun	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld) 2.03
0	TIME OF COUNT	(TOC)	14:23	Sr-89/90 CONC in µCi/g	3.9558E+00
Sample Prep	DATE OF COUNT	(DOC)	12/07/98		
FUSION01					
Sample#					
S98T002739					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date					
12/07/98	Sr-89/90 CONCENTRATION		3.96E+00	µCi/g	DETECTION LEVEL
Analysis Time					
12:25 PM	RELATIVE COUNTING ERROR		1.0%		2.26E-03
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY		91.7%		µCi/g

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	8 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4721

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	64934	GROSS WEIGHT	(W2) 7.3924
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3003
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3) 0.0921
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 4.65
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9852		
SOLID	SAMPLE COUNT RATE	(Rs)	6490.80	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005753	TIME OF SEPARATION	(ST)	09:55	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld) 2.03
0	TIME OF COUNT	(TOC)	14:34	Sr-89/90 CONC in $\mu\text{Ci/g}$	3.6276E+01
Sample Prep	DATE OF COUNT	(DOC)	12/07/98		
FUSION01					
Sample #					
S98T002844					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date					DETECTION LEVEL 1.14E-02 $\mu\text{Ci/g}$
12/07/98	Sr-89/90 CONCENTRATION	3.63E+01	$\mu\text{Ci/g}$		
Analysis Time					
12:25 PM	RELATIVE COUNTING ERROR	0.8%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	92.1%			

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	8 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4722

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	65609	GROSS WEIGHT	(W2) 7.4083
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3166
27136	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3) 0.0917
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 4.83
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9944		
SOLID	SAMPLE COUNT RATE	(Rs)	6558.30	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005753	TIME OF SEPARATION	(ST)	09:55	Rmax	N/A
Rein	DATE OF SEPARATION	(SD)	12/07/98	DETECTION LIMIT	(Ld) 2.03
0	TIME OF COUNT	(TOC)	14:45	Sr-89/90 CONC in $\mu\text{Ci/g}$	3.6571E+01
Sample Prep	DATE OF COUNT	(DOC)	12/07/98		
FUSION01					
Sample #					
S98T002844					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)				
VAR	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/07/98					
Analysis Date					DETECTION LEVEL 1.13E-02 $\mu\text{Ci/g}$
12/07/98	Sr-89/90 CONCENTRATION	3.66E+01	$\mu\text{Ci/g}$		
Analysis Time					
12:25 PM	RELATIVE COUNTING ERROR	0.8%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	91.7%			

Analyst:	GLL	Date:	07-Dec-98
Signature of Chemist:	SAC	Date:	8 Dec 98

SAMPLE.WB1 REV 2.0

22010NML

4723

LABCORE Completed Worklist Report for Worklist# 27137

Analyst: gl Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY105, @SR90-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@SR90-01 SR90-01 SOLID		1.04E-03	9.82E-4	94.423 % Recovery	
1 STD	0		@SR90-01 SR90-01C SOLID		100	9.38E+01	93.800 % Recovery	
1 STD	0		@SR90-01 SR90-01E SOLID		1.00	2.01E+00	2.010 % Ct. Err	
2 BLNK-PREP	0		@SR90-01 SR90-01 SOLID		1	7.35E-3	0.007 uCi/g	
2 BLNK-PREP	0		@SR90-01 SR90-01C SOLID		100	9.25E+01	92.500 % Recovery	
2 BLNK-PREP	0		@SR90-01 SR90-01E SOLID		1.00	4.81E+01	48.100 uCi/g	
3 BLNK/BKG	0		@SR90-01 SR90-01 SOLID		1	2.71E+00	2.710 BLNK/BKG	
4 SAMPLE	S98T002920	0 F	@SR90-01 SR90-01 SOLID	N/A		5.96E-01	3.78E-003 uCi/g	
4 SAMPLE	S98T002920	0 F	@SR90-01 SR90-01C SOLID	N/A		9.16E+01	% Recovery	
4 SAMPLE	S98T002920	0 F	@SR90-01 SR90-01E SOLID	N/A		3.67E+00	% Ct. Error	
5 DUP	S98T002920	0 F	@SR90-01 SR90-01 SOLID		5.96E-1	6.90E-1	14.619 RPD	
5 DUP	S98T002920	0 F	@SR90-01 SR90-01C SOLID		100	9.18E+01	91.800 % Recovery	
5 DUP	S98T002920	0 F	@SR90-01 SR90-01E SOLID		1.00	3.40E+00	3.400 % Ct. Err	
6 SAMPLE	S98T002921	0 F	@SR90-01 SR90-01 SOLID	N/A		6.43E-01	3.71E-003 uCi/g	
6 SAMPLE	S98T002921	0 F	@SR90-01 SR90-01C SOLID	N/A		9.32E+01	% Recovery	
6 SAMPLE	S98T002921	0 F	@SR90-01 SR90-01E SOLID	N/A		3.50E+00	% Ct. Error	
7 DUP	S98T002921	0 F	@SR90-01 SR90-01 SOLID		6.43E-1	7.47E-1	14.964 RPD	
7 DUP	S98T002921	0 F	@SR90-01 SR90-01C SOLID		100	9.32E+01	93.200 % Recovery	
7 DUP	S98T002921	0 F	@SR90-01 SR90-01E SOLID		1.00	3.25E+00	3.250 % Ct. Err	
8 SAMPLE	S98T002922	0 F	@SR90-01 SR90-01 SOLID	N/A		3.42E+02	9.73E-002 uCi/g	
8 SAMPLE	S98T002922	0 F	@SR90-01 SR90-01C SOLID	N/A		9.29E+01	% Recovery	
8 SAMPLE	S98T002922	0 F	@SR90-01 SR90-01E SOLID	N/A		7.72E-01	% Ct. Error	
9 DUP	S98T002922	0 F	@SR90-01 SR90-01 SOLID		3.42E+2	1.59E+2	73.054 RPD	
9 DUP	S98T002922	0 F	@SR90-01 SR90-01C SOLID		100	9.31E+01	93.100 % Recovery	
9 DUP	S98T002922	0 F	@SR90-01 SR90-01E SOLID		1.00	1.13E+00	1.130 % Ct. Err	

Final page for worklist# 27137

Analyst Signature _____ Date _____ Analyst Signature _____ Date _____

[Signature] 25 Nov 98
Reviewer Signature Date

High RPD for sample S98T002922, most likely due to sample inhomogeneity. Dose rates are much different also. Results accepted per PC doc.

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27137

Analyst: gl Instrument: AB00 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105, @SR90-01, tdm

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@SR90-01	SOLID		
2	BLNK-PREP				@SR90-01	SOLID		
3	BLNK/BKG				@SR90-01	SOLID		
4	SAMPLE	S98T002920	0	F	@SR90-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: SR90-01, SR90-01C, SR90-01E						
5	DUP	S98T002920	0	F	@SR90-01	SOLID		
6	SAMPLE	S98T002921	0	F	@SR90-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: SR90-01, SR90-01C, SR90-01E						
7	DUP	S98T002921	0	F	@SR90-01	SOLID		
8	SAMPLE	S98T002922	0	F	@SR90-01	SOLID	98000490	BY-105 (2)
		Analytes Requested: SR90-01, SR90-01C, SR90-01E						
9	DUP	S98T002922	0	F	@SR90-01	SOLID		

Final page for worklist # 27137

Dee L. H. A. 11-24-98
Analyst Signature Date

Sharon L. H. A. 11-25-98
Analyst Signature Date
MW 11/25/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	9560	GROSS WEIGHT (W2)	7.4137
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3199
27137	BACKGROUND in cpm (BKG)	2.1	NET WEIGHT (W3)	0.0938
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	10.17
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR DDF	1		
LIQUID	SAMPLE COUNT RATE (Rs)	953.90	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.87	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005754	TIME OF SEPARATION (ST)	10:50	Rmax	N/A
Rerun	DATE OF SEPARATION (SD)	11/24/98	DETECTION LIMIT (Ld)	1.84
0	TIME OF COUNT (TOC)	21:00	Sr-89/90 CONC. in $\mu\text{Ci/L}$	9.8208E-01
Sample Prep	DATE OF COUNT (DOC)	11/24/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0423E-03		

WL27137-STD

Instrument Code	WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)		
Prepared By	SLH2	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $\text{RS} \leq \text{Lc}$ AND $\text{RS} \geq 0$ OR REPLACE RS WITH Lc IF $\text{RS} < 0$		
Chemist	SAC	$\text{RS} \cdot \text{DF} \cdot \text{DDF} \cdot 1000 / ((\text{C1} + \text{C2}) \cdot (1 - e^{-(\ln(2) \cdot \text{DT})})) \cdot \text{SS} \cdot \text{REC} \cdot 2220000$ NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} \cdot 0.1000))$		
Analyst	GLL	Relative Counting Error = The Square Root of $((\text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT}) \cdot 1.96)$ Percent Carrier Recovery = (Net Weight / Expected weight) * 100 NOTE: Expected weight = $\text{CVA} \cdot 0.1$		
Date Complete	11/25/98	Detection Levels and Less Than Values are determined from Procedure LA-508-002. Delta Time (hours) = $((\text{DOC} - \text{SD}) \cdot 24) + (\text{TOC} - \text{ST}) / 100$		
Analysis Date	11/24/98	Sr-89/90 CONCENTRATION	9.82E-04	DETECTION LEVEL 1.89E-06 $\mu\text{Ci/L}$
Analysis Time	02:40 PM	RELATIVE COUNTING ERROR	2.0%	
Sample Point				
BY-105 (2)		PERCENT CARRIER RECOVERY	93.8%	

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	<i>[Signature]</i>	Date:	25 Nov 98

STANDARD.WB1 REV 2.0

22010NML

4726

HNF-1663 REV. 0

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

LA-220-101 / E-3				Sr-89/90 LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)		BLNK-PREP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000	
BLNK-PREP	TOTAL COUNTS	(TC)	57	GROSS WEIGHT	(W2)	7.3887	
Worksheet	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.2962	
27137	BACKGROUND in cpm	(BKG)	2.1	NET WEIGHT	(W3)	0.0925	
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	10.33	
@SR90-01	DILUTION FACTOR	DF	1				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0416				
SOLID	SAMPLE COUNT RATE	(Rs)	3.60	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005754	TIME OF SEPARATION	(ST)	10:50	Rmax		N/A	
Return	DATE OF SEPARATION	(SD)	11/24/98	DETECTION LIMIT	(Ld)	1.84	
0	TIME OF COUNT	(TOC)	21:10	Sr-89/90 CONC in µCi/g			7.3519E-03
Sample Prep	DATE OF COUNT	(DOC)	11/24/98				
N/A							
Sample #							
WL27137-BLNK							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/25/98							
Analysis Date					DETECTION LEVEL 3.75E-03 µCi/g		
11/24/98	Sr-89/90 CONCENTRATION		7.35E-03	µCi/g			
Analysis Time							
02:40 PM	RELATIVE COUNTING ERROR		48.1%				
Sample Point							
BY-105 (2)	PERCENT CARRIER RECOVERY		92.5%				

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	SAC	Date:	25 Nov 98

BLANK.WB1 REV 2.0

22010NML

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	2916	GROSS WEIGHT	(W2)	7.3966
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3050
27137	BACKGROUND in cpm (BKG)	2.1	NET WEIGHT	(W3)	0.0916
Test Code	SAMPLE VOLUME in mL (SS)	0.250	DELTA TIME (HOURS)	(DT)	10.50
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST FACTOR (g/L) (D g/L)	2.0416			
SOLID	SAMPLE COUNT RATE (Rs)	289.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005754	TIME OF SEPARATION (ST)	10:50	Rmax		N/A
Re-run	DATE OF SEPARATION (SD)	11/24/98	DETECTION LIMIT	(Ld)	1.84
0	TIME OF COUNT (TOC)	21:20	Sr-89/90 CONC in $\mu\text{Ci/g}$		5.9607E-01
Sample Prep	DATE OF COUNT (DOC)	11/24/98			
FUSION01					
Sample #					
S98T002920					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*22200000)				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/25/98					
Analysis Date					DETECTION LEVEL 3.78E-03 $\mu\text{Ci/g}$
11/24/98	Sr-89/90 CONCENTRATION	5.96E-01	$\mu\text{Ci/g}$		
Analysis Time					
02:40 PM	RELATIVE COUNTING ERROR	3.7%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	91.6%			

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	<i>[Signature]</i>	Date:	25 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4728

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS (TC)	3382	GROSS WEIGHT	(W2)	7.3986
Work Day	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3068
27137	BACKGROUND in cpm (BKG)	2.1	NET WEIGHT	(W3)	0.0918
Test Code	SAMPLE VOLUME in mL (SS)	0.250	DELTA TIME (HOURS)	(DT)	10.67
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST FACTOR (g/L) (D g/L)	2.0408			
SOLID	SAMPLE COUNT RATE (Rs)	336.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005754	TIME OF SEPARATION (ST)	10:50	Rmax		N/A
Refun	DATE OF SEPARATION (SD)	11/24/98	DETECTION LIMIT	(Ld)	1.84
0	TIME OF COUNT (TOC)	21:30	Sr-89/90 CONC in $\mu\text{Ci/g}$		6.8968E-01
Sample Prep	DATE OF COUNT (DOC)	11/24/98			
FUSION01					
Sample #					
S98T002920					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} * 1000 * \text{DF} / ((\text{C1} + \text{C2} * (1 - e^{(-\text{natural log } 2 / 64.2 * \text{DT}))}) * \text{SS} * (\text{Dg/L} * \text{REC} * 2220000))$				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$				
11/25/98					
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL
11/24/98	6.90E-01 $\mu\text{Ci/g}$				
Analysis Time	RELATIVE COUNTING ERROR				
02:40 PM	3.4%				
Sample Point					3.77E-03 $\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY				91.8%

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	<i>[Signature]</i>	SAC	Date: 25 Nov 98
SAMPLE.WB1 REV 2.0 22010NML			

4729

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
TYPE	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	3201	GROSS WEIGHT (W2)		7.3987
Worklist	COUNT TIME IN MINUTES (CT)	10	TARE WEIGHT (W1)		7.3055
27137	BACKGROUND in cpm (BKG)	2.1	NET WEIGHT (W3)		0.0932
Test Code	SAMPLE VOLUME in mL (SS)	0.250	DELTA TIME (HOURS) (DT)		10.83
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST FACTOR (g/L) (D g/L)	2.038			
SOLID	SAMPLE COUNT RATE (Rs)	318.00	SR-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.87	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98005754	TIME OF SEPARATION (ST)	10:50	Rmax		N/A
Re-run	DATE OF SEPARATION (SD)	11/24/98	DETECTION LIMIT (Ld)		1.84
0	TIME OF COUNT (TOC)	21:40	Sr-89/90 CONC in $\mu\text{Ci/g}$		6.4259E-01
Sample Prep	DATE OF COUNT (DOC)	11/24/98			
FUSION01					
Sample #					
S98T002921					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS <= Lc and RS >= 0 or Replace RS with Lc if RS < 0				
Prepared By	RS * 1000 * DF / ((C1 + C2 * (1 - e to the power of ((-natural log 2) / (64.2 * DT)))) * SS * (Dg/L) * REC * 2220000)				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/25/98					
Analysis Date					DETECTION LEVEL
11/24/98	Sr-89/90 CONCENTRATION	6.43E-01	$\mu\text{Ci/g}$		
Analysis Time					3.71E-03 $\mu\text{Ci/g}$
02:40 PM	RELATIVE COUNTING ERROR	3.5%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	93.2%			

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	<i>SLCattus</i>	Date:	25 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4730

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	3701	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3468
27137	BACKGROUND in cpm	(BKG)	2.1	NET WEIGHT	(W3)	0.0932
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	11.08
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.024			
SOLID	SAMPLE COUNT RATE	(Rs)	368.00	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005754	TIME OF SEPARATION	(ST)	10:50	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/24/98	DETECTION LIMIT	(Ld)	1.84
0	TIME OF COUNT	(TOC)	21:55	Sr-89/90 CONC in $\mu\text{Ci/g}$		7.4699E-01
Sample Prep	DATE OF COUNT	(DOC)	11/24/98			
FUSION01						
Sample #						
S98T002921						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/25/98						
Analysis Date						DETECTION LEVEL 3.73E-03 $\mu\text{Ci/g}$
11/24/98	Sr-89/90 CONCENTRATION		7.47E-01	$\mu\text{Ci/g}$		
Analysis Time						
02:40 PM	RELATIVE COUNTING ERROR		3.2%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		93.2%			

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	SAC	Date:	25 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4731

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000	SAMPLE
SAMPLE	TOTAL COUNTS	(TC)	64563	GROSS WEIGHT	(W2)	7.4077	
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3148	
27137	BACKGROUND in cpm	(BKG)	2.1	NET WEIGHT	(W3)	0.0929	
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	11.33	
@SR90-01	DILUTION FACTOR	DF	101				
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9608				
SOLID	SAMPLE COUNT RATE	(Rs)	6454.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98005754	TIME OF SEPARATION	(ST)	10:50	Rmax		N/A	
Return	DATE OF SEPARATION	(SD)	11/24/98	DETECTION LIMIT	(Ld)	1.84	
0	TIME OF COUNT	(TOC)	22:10	Sr-89/90 CONC in $\mu\text{Ci/g}$		3.4176E+02	
Sample Prep	DATE OF COUNT	(DOC)	11/24/98				
FUSION01							
Sample #							
S98T002922							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0						
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)						
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Completed	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
11/25/98							
Analysis Date							
11/24/98	Sr-89/90 CONCENTRATION		3.42E+02		$\mu\text{Ci/g}$		DETECTION LEVEL
Analysis time							
02:40 PM	RELATIVE COUNTING ERROR		0.8%				9.73E-02
Sample Point							$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY		92.9%				

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	SAC	Date:	25 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4732

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	30251	GROSS WEIGHT	(W2) 7.1366
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.0435
27137	BACKGROUND in cpm	(BKG)	2.1	NET WEIGHT	(W3) 0.0931
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 11.58
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9668		
SOLID	SAMPLE COUNT RATE	(Rs)	3023.00	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.87	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005754	TIME OF SEPARATION	(ST)	10:50	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	11/24/98	DETECTION LIMIT	(Ld) 1.84
0	TIME OF COUNT	(TOC)	22:25	Sr-89/90 CONC in µCi/g	1.5887E+02
Sample Prep	DATE OF COUNT	(DOC)	11/24/98		
FUSION01					
Sample #					
S98T002922					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/25/98					
Analysis Date					DETECTION LEVEL 9.65E-02 µCi/g
11/24/98	Sr-89/90 CONCENTRATION	1.59E+02		µCi/g	
Analysis Time					
02:40 PM	RELATIVE COUNTING ERROR	1.1%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	93.1%			

Analyst:	GLL	Date:	25-Nov-98
Signature of Chemist:	SAC	Date:	25 Nov 98
SAMPLE.WB1 REV 2.0 22010NML			

4733

LABCORE Completed Worklist Report for Worklist# 27138

Analyst: gll Instrument: AB12 Book#: _____
Method: LA-220-101 Rev/Mod _____
Worklist Comment: BY105, @SR90-01, tdm

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01	SR90-01 SOLID		1.04E+03	1.04E+3	100.000 %	Recovery
1 STD	0	@SR90-01	SR90-01C SOLID		100	8.48E+01	84.800 %	Recovery
1 STD	0	@SR90-01	SR90-01E SOLID		1.00	2.12E+00	2.120 %	Ct. Err
2 BLNK-PREP	0	@SR90-01	SR90-01 SOLID		1	9.64E+3	0.010 uCi/g	
2 BLNK-PREP	0	@SR90-01	SR90-01C SOLID		100	8.40E+01	84.000 %	Recovery
2 BLNK-PREP	0	@SR90-01	SR90-01E SOLID		1.00	4.74E+01	47.400 uCi/g	
3 BLNK/BKG	0	@SR90-01	SR90-01 SOLID		1	2.56E+00	2.560 BLNK/BKG	
4 SAMPLE	S98T002986	0 F	@SR90-01	SR90-01 SOLID	N/A	3.81E+00	5.27e-003 uCi/g	
4 SAMPLE	S98T002986	0 F	@SR90-01	SR90-01C SOLID	N/A	7.84E+01	% Recovery	
4 SAMPLE	S98T002986	0 F	@SR90-01	SR90-01E SOLID	N/A	1.63E+00	% Ct. Error	
5 DUP	S98T002986	0 F	@SR90-01	SR90-01 SOLID	3.81E+0	4.43E+0	15.049 RPD	
5 DUP	S98T002986	0 F	@SR90-01	SR90-01C SOLID	100	8.12E+01	81.200 %	Recovery
5 DUP	S98T002986	0 F	@SR90-01	SR90-01E SOLID	1.00	1.48E+00	1.480 %	Cnt Err
6 SAMPLE	S98T002987	0 F	@SR90-01	SR90-01 SOLID	N/A	6.20E+00	4.64e-003 uCi/g	
6 SAMPLE	S98T002987	0 F	@SR90-01	SR90-01C SOLID	N/A	8.58E+01	% Recovery	
6 SAMPLE	S98T002987	0 F	@SR90-01	SR90-01E SOLID	N/A	1.20E+00	% Ct. Error	
7 DUP	S98T002987	0 F	@SR90-01	SR90-01 SOLID	6.20E+0	6.68E+0	7.453 RPD	
7 DUP	S98T002987	0 F	@SR90-01	SR90-01C SOLID	100	8.21E+01	82.100 %	Recovery
7 DUP	S98T002987	0 F	@SR90-01	SR90-01E SOLID	1.00	1.18E+00	1.180 %	Cnt Err

Final page for worklist# 27138

Analyst Signature	Date	Analyst Signature	Date
Reviewer Signature	Date		
	2/7/98		

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27138

Analyst: gll Instrument: AB00 12 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105, @SR90-01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002986 0 F	@SR90-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: SR90-01, SR90-01C, SR90-01E					
5 DUP	S98T002986 0 F	@SR90-01	SOLID			
6 SAMPLE	S98T002987 0 F	@SR90-01	SOLID		98000509	BY-105 (2)
	Analytes Requested: SR90-01, SR90-01C, SR90-01E					
7 DUP	S98T002987 0 F	@SR90-01	SOLID			

Final page for worklist # 27138

[Signature] 11-20-98
Analyst Signature Date

[Signature] 11/23/98
Analyst Signature Date
[Signature] 11/20/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	8591	GROSS WEIGHT (W2)	7.8690
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.7842
27138	BACKGROUND in cpm (BKG)	2.5	NET WEIGHT (W3)	0.0848
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	4.22
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR (DDF)	1		
LIQUID	SAMPLE COUNT RATE (Rs)	856.60	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.95	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005755	TIME OF SEPARATION (ST)	20:50	Rmax	N/A
Re-run	DATE OF SEPARATION (SD)	11/20/98	DETECTION LIMIT (Ld)	2.00
0	TIME OF COUNT (TOC)	01:03	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0371E+00
Sample Prep	DATE OF COUNT (DOC)	11/21/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0426E-03		
WL27138				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS > 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
SZC	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((-\text{natural log } 2) / 64.2 * DT))) * SS * REC * 2220000)$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
GLL	NOTE: Expected weight = $CVA * 0.1$			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
11/23/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date				
11/20/98	Sr-89/90 CONCENTRATION	1.04E-03	$\mu\text{Ci/mL}$	DETECTION LEVEL
Analysis Time				
11:30 PM	RELATIVE COUNTING ERROR	2.1%		2.42E-06
Sample Point				$\mu\text{Ci/L}$
BY-105 (2)	PERCENT CARRIER RECOVERY	84.8%		

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:		SAC	Date: 24 Nov 98

STANDARD WB1 REV 2.0

22010NML

4736

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	64	GROSS WEIGHT	(W2)	7.9018
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.8178
27138	BACKGROUND in cpm	(BKG)	2.5	NET WEIGHT	(W3)	0.0840
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	4.53
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.972			
SOLID	SAMPLE COUNT RATE	(Rs)	3.90	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.95	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005755	TIME OF SEPARATION	(ST)	20:50	Rmax		N/A
Retun	DATE OF SEPARATION	(SD)	11/20/98	DETECTION LIMIT	(Ld)	2.00
0	TIME OF COUNT	(TOC)	01:22	Sr-89/90 CONC in $\mu\text{Ci/g}$		9.6356E-03
Sample Prep	DATE OF COUNT	(DOC)	11/21/98			
N/A						
Sample #						
WL27138						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/23/98						
Analysis Date						DETECTION LEVEL 4.93E-03 $\mu\text{Ci/g}$
11/20/98	Sr-89/90 CONCENTRATION		9.64E-03	$\mu\text{Ci/g}$		
Analysis Time						
11:30 PM	RELATIVE COUNTING ERROR		47.4%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		84.0%			

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:	SAC	Date:	24 Nov 98

BLANK.WB1 REV 2.0

22010NML

4737

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	14462	GROSS WEIGHT	(W2) 7.8629
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7845
27138	BACKGROUND in cpm	(BKG)	2.5	NET WEIGHT	(W3) 0.0784
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 4.83
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.972		
SOLID	SAMPLE COUNT RATE	(Rs)	1443.70	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.95	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005755	TIME OF SEPARATION	(ST)	20:50	Rmax	N/A
Refund	DATE OF SEPARATION	(SD)	11/20/98	DETECTION LIMIT	(Ld) 2.00
0	TIME OF COUNT	(TOC)	01:40	Sr-89/90 CONC in $\mu\text{Ci/g}$	3.8092E+00
Sample Prep	DATE OF COUNT	(DOC)	11/21/98		
FUSION01					
Sample #					
S98T002986					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/23/98					
Analysis Date					DETECTION LEVEL
11/20/98	Sr-89/90 CONCENTRATION	3.81E+00	$\mu\text{Ci/g}$		
Analysis Time					
11:30 PM	RELATIVE COUNTING ERROR	1.6%			5.27E-03
Sample Point					$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY	78.4%			

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:	SAC	Date:	24 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4738

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	17617	GROSS WEIGHT	(W2) 7.8860
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.8048
27138	BACKGROUND in cpm	(BKG)	2.5	NET WEIGHT	(W3) 0.0812
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 5.30
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.984		
SOLID	SAMPLE COUNT RATE	(Rs)	1759.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.95	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005755	TIME OF SEPARATION	(ST)	20:50	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	11/20/98	DETECTION LIMIT	(Ld) 2.00
0	TIME OF COUNT	(TOC)	02:08	Sr-89/90 CONC in $\mu\text{Ci/g}$	4.4322E+00
Sample Prep	DATE OF COUNT	(DOC)	11/21/98		
FUSION01					
Sample #					
S98T002986					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/23/98					
Analysis Date					DETECTION LEVEL 5.03E-03 $\mu\text{Ci/g}$
11/20/98	Sr-89/90 CONCENTRATION	4.43E+00	$\mu\text{Ci/g}$		
Analysis Time					
11:30 PM	RELATIVE COUNTING ERROR	1.5%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	81.2%			

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:	SAC	Date:	24 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4729

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	26722	GROSS WEIGHT	(W2) 7.8837
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7979
27138	BACKGROUND in cpm	(BKG)	2.5	NET WEIGHT	(W3) 0.0858
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 5.75
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0272		
SOLID	SAMPLE COUNT RATE	(Rs)	2669.70	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.95	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005755	TIME OF SEPARATION	(ST)	20:50	Rmax	N/A
Retain	DATE OF SEPARATION	(SD)	11/20/98	DETECTION LIMIT	(Ld) 2.00
0	TIME OF COUNT	(TOC)	02:35	Sr-89/90 CONC in $\mu\text{Ci/g}$	6.2001E+00
Sample Prep	DATE OF COUNT	(DOC)	11/21/98		
FUSION01					
Sample #					
S98T002987					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/23/98					
Analysis Date					DETECTION LEVEL 4.64E-03 $\mu\text{Ci/g}$
11/20/98	Sr-89/90 CONCENTRATION	6.20E+00	$\mu\text{Ci/g}$		
Analysis Time					
11:30 PM	RELATIVE COUNTING ERROR	1.2%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	85.8%			

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:	SAC	Date:	24 Nov 98
SAMPLE.WB1 REV 2.0 22010NML			

4740

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS (TC)	27656	GROSS WEIGHT	(W2)	7.8570
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.7749
27138	BACKGROUND in cpm (BKG)	2.5	NET WEIGHT	(W3)	0.0821
Test Code	SAMPLE VOLUME in mL (SS)	0.250	DELTA TIME (HOURS)	(DT)	5.95
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST FACTOR (g/L) (D g/L)	2.0312			
SOLID	SAMPLE COUNT RATE (Rs)	2763.10	Sr-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.95	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98005755	TIME OF SEPARATION (ST)	20:50	Rmax		N/A
Retun	DATE OF SEPARATION (SD)	11/20/98	DETECTION LIMIT	(Ld)	2.00
0	TIME OF COUNT (TOC)	02:47	Sr-89/90 CONC in $\mu\text{Ci/g}$ 6.6789E+00		
Sample Prep	DATE OF COUNT (DOC)	11/21/98			
FUSION01					
Sample #					
S98T002987					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared by	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/23/98					
Analysis Date	Sr-89/90 CONCENTRATION 6.68E+00 $\mu\text{Ci/g}$				DETECTION LEVEL 4.82E-03 $\mu\text{Ci/g}$
11/20/98					
Analysis Time	RELATIVE COUNTING ERROR 1.2%				
11:30 PM					
Sample Point	PERCENT CARRIER RECOVERY 82.1%				
BY-105 (2)					

Analyst:	GLL	Date:	23-Nov-98
Signature of Chemist:	SAC	Date:	24 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4741

LABCORE Completed Worklist Report for Worklist# 27139

Analyst: crj Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY105, @SR90-01, tdm

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 SOLID		1.04E-03	1.03E-3	99.03% Recovery	
1 STD	0	@SR90-01 SR90-01C SOLID		100	9.29E+01	92.90% Recovery	
1 STD	0	@SR90-01 SR90-01E SOLID		1.00	2.05E+00	2.05% Ct. Err	
2 BLNK-PREP	0	@SR90-01 SR90-01 SOLID		1	7.17E-3	0.007 uCi/g	
2 BLNK-PREP	0	@SR90-01 SR90-01C SOLID		100	9.32E+01	93.20% Recovery	
2 BLNK-PREP	0	@SR90-01 SR90-01E SOLID		1.00	6.06E+01	60.600 uCi/g	
3 BLNK/BKG	0	@SR90-01 SR90-01 SOLID		1	1.97E+00	1.970 BLNK/BKG	
4 SAMPLE	S98T002988 0 F	@SR90-01 SR90-01 SOLID	N/A	4.77E+00	5.12E-003	uCi/g	
4 SAMPLE	S98T002988 0 F	@SR90-01 SR90-01C SOLID	N/A	9.29E+01		% Recovery	
4 SAMPLE	S98T002988 0 F	@SR90-01 SR90-01E SOLID	N/A	1.35E+00		% Ct. Error	
5 DUP	S98T002988 0 F	@SR90-01 SR90-01 SOLID		4.77E+0	3.49E+0	30.953 RPD	
5 DUP	S98T002988 0 F	@SR90-01 SR90-01C SOLID		100	9.35E+01	93.500 % Recovery	
5 DUP	S98T002988 0 F	@SR90-01 SR90-01E SOLID		1.00	1.56E+00	1.560 % Ct. Err	
6 SAMPLE	S98T002989 0 F	@SR90-01 SR90-01 SOLID	N/A	3.87E+00	4.97E-003	uCi/g	
6 SAMPLE	S98T002989 0 F	@SR90-01 SR90-01C SOLID	N/A	9.26E+01		% Recovery	
6 SAMPLE	S98T002989 0 F	@SR90-01 SR90-01E SOLID	N/A	1.47E+00		% Ct. Error	
7 DUP	S98T002989 0 F	@SR90-01 SR90-01 SOLID		3.87E+0	3.76E+0	2.883 RPD	
7 DUP	S98T002989 0 F	@SR90-01 SR90-01C SOLID		100	9.27E+01	92.700 % Recovery	
7 DUP	S98T002989 0 F	@SR90-01 SR90-01E SOLID		1.00	1.49E+00	1.490 % Ct. Err	

Final page for worklist# 27139

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature Date

Sample S98T002988 will be run dist RPD re
Reanalyzed on WL# 27743.
3/4/99

LABCORE Data Entry Template for Worklist# 27139

Analyst: CRJ Instrument: AB00 12 Book# 28B57

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105, @SR90-01, tdm

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002988	0 F	@SR90-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002988	0 F	@SR90-01	SOLID		
6 SAMPLE	S98T002989	0 F	@SR90-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
7 DUP	S98T002989	0 F	@SR90-01	SOLID		

Final page for worklist # 27139

CRJ 12/29/98
Analyst Signature Date

Emil Buech 12/30/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	9220	GROSS WEIGHT (W2)	7.4038
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3109
27139	BACKGROUND in cpm (BKG)	3.3	NET WEIGHT (W3)	0.0929
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	3.10
@SR90-01	DILUTION FACTOR (DF)	1		
Matrx	DIGEST DILUTION FACTOR DDF	1		
LIQUID	SAMPLE COUNT RATE (Rs)	918.70	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	1.09	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005756	TIME OF SEPARATION (ST)	13:14	Rmax	N/A
Rerun	DATE OF SEPARATION (SD)	12/29/98	DETECTION LIMIT (Ld)	2.28
0	TIME OF COUNT (TOC)	16:20	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0280E+00
Sample Prep	DATE OF COUNT (DOC)	12/29/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0399E-03		
WL27139-STD				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS \geq 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
EMB	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((-\text{natural log } 2) / 64.2 * DT)))) * SS * REC * 2220000$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
CRJ	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
12/30/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date	Sr-89/90 CONCENTRATION			DETECTION LEVEL 2.55E-06 $\mu\text{Ci/L}$
12/29/98	1.03E-03 $\mu\text{Ci/mL}$			
Analysis Time	RELATIVE COUNTING ERROR			
03:30 PM	2.1%			
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY			92.9%

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	SAC	Date:	

STANDARD.WB1 REV 2.0

22010NML

4744

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	65	GROSS WEIGHT	(W2)	7.4252
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3320
27139	BACKGROUND in cpm	(BKG)	3.3	NET WEIGHT	(W3)	0.0932
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	3.27
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9872			
SOLID	SAMPLE COUNT RATE	(Rs)	3.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.09	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005756	TIME OF SEPARATION	(ST)	13:14	Rmax		N/A
Re-run	DATE OF SEPARATION	(SD)	12/29/98	DETECTION LIMIT	(Ld)	2.28
0	TIME OF COUNT	(TOC)	16:30	Sr-89/90 CONC in $\mu\text{Ci/g}$		7.1707E-03
Sample Prep	DATE OF COUNT	(DOC)	12/29/98			
N/A						
Sample #						

WL27139 BLNK-PREP

Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					DETECTION LEVEL 5.11E-03 μCi/g
WB27811	Sr-89/90 CONC in μCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT)))*SS*(Dg/L)*REC*22200000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/30/98						
Analysis Date						
12/29/98	Sr-89/90 CONCENTRATION	7.17E-03	μCi/g			
Analysis Time						
03:30 PM	RELATIVE COUNTING ERROR	60.6%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	93.2%				

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	<i>[Signature]</i>	Date:	<i>30 Dec 98</i>

BLANK.WB1 REV 2.0

22010NML

4745

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	21280	GROSS WEIGHT	(W2) 7.4186
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3257
27139	BACKGROUND in cpm	(BKG)	3.3	NET WEIGHT	(W3) 0.0929
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT) 3.43
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9872		
SOLID	SAMPLE COUNT RATE	(Rs)	2124.70	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.09	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005756	TIME OF SEPARATION	(ST)	13:14	Rmax	N/A
ReRun	DATE OF SEPARATION	(SD)	12/29/98	DETECTION LIMIT	(Ld) 2.28
0	TIME OF COUNT	(TOC)	16:40	Sr-89/90 CONC in $\mu\text{Ci/g}$	4.7676E+00
Sample Prep	DATE OF COUNT	(DOC)	12/29/98		
FUSION01					
Sample #					
S98T002988					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/30/98					
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL 5.12E-03 $\mu\text{Ci/g}$
12/29/98	4.77E+00 $\mu\text{Ci/g}$				
Analysis Time	RELATIVE COUNTING ERROR				
03:30 PM	1.3%				
Sample Point	PERCENT CARRIER RECOVERY				
BY-105 (2)	92.9%				

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	SAC	Date:	30 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

DATA
NOT USED

4746

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	15830	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3213
27139	BACKGROUND in cpm	(BKG)	3.3	NET WEIGHT	(W3)	0.0935
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	3.60
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9992			
SOLID	SAMPLE COUNT RATE	(Rs)	1579.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.09	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005756	TIME OF SEPARATION	(ST)	13:14	Rmax		N/A
Run	DATE OF SEPARATION	(SD)	12/29/98	DETECTION LIMIT	(Ld)	2.28
0	TIME OF COUNT	(TOC)	16:50	Sr-89/90 CONC in µCi/g		3.4943E+00
Sample Prep	DATE OF COUNT	(DOC)	12/29/98			
FUSION01						
Sample #						
S98T002988						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/30/98						
Analysis Date	Sr-89/90 CONCENTRATION		3.49E+00		µCi/g	DETECTION LEVEL 5.04E-03 µCi/g
Analysis Time	RELATIVE COUNTING ERROR		1.6%			
03:30 PM						
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		93.5%			

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	SAC	Date:	30 Dec 98
SAMPLE.WB1 REV 2.0	22010NML		

DATA
NOT USED

4747

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	17798	GROSS WEIGHT	(W2)	7.4156
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3230
27139	BACKGROUND in cpm	(BKG)	3.3	NET WEIGHT	(W3)	0.0926
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	4.93
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0192			
SOLID	SAMPLE COUNT RATE	(Rs)	1776.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.09	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005756	TIME OF SEPARATION	(ST)	13:19	Rmax		N/A
SR Run	DATE OF SEPARATION	(SD)	12/29/98	DETECTION LIMIT	(Ld)	2.28
0	TIME OF COUNT	(TOC)	18:15	Sr-89/90 CONC in $\mu\text{Ci/g}$		3.8716E+00
Sample Prep	DATE OF COUNT	(DOC)	12/29/98			
FUSION01						
Sample #						
S98T002989						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/30/98						
Analysis Date						DETECTION LEVEL 4.97E-03 $\mu\text{Ci/g}$
12/29/98	Sr-89/90 CONCENTRATION	3.87E+00		$\mu\text{Ci/g}$		
Analysis Time						
03:30 PM	RELATIVE COUNTING ERROR	1.5%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	92.6%				

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	SAC	Date:	30 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4748

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	17383	GROSS WEIGHT	(W2)	7.4339
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3412
27139	BACKGROUND in cpm	(BKG)	3.3	NET WEIGHT	(W3)	0.0927
Test Code	SAMPLE VOLUME in mL	(SS)	0.250	DELTA TIME (HOURS)	(DT)	5.18
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.0228			
SOLID	SAMPLE COUNT RATE	(Rs)	1735.00	Sr-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.09	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005756	TIME OF SEPARATION	(ST)	13:19	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/29/98	DETECTION LIMIT	(Ld)	2.28
0	TIME OF COUNT	(TOC)	18:30	Sr-89/90 CONC in $\mu\text{Ci/g}$		3.7602E+00
Sample Prep	DATE OF COUNT	(DOC)	12/29/98			
FUSION01						
Sample #						
S98T002989						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/30/98						
Analysis Date						DETECTION LEVEL
12/29/98	Sr-89/90 CONCENTRATION		3.76E+00		$\mu\text{Ci/g}$	
Analysis Time						
03:30 PM	RELATIVE COUNTING ERROR		1.5%			
Sample Point						4.94E-03
BY-105 (2)	PERCENT CARRIER RECOVERY		92.7%			$\mu\text{Ci/g}$

Analyst:	CRJ	Date:	30-Dec-98
Signature of Chemist:	SAC	Date:	30-Dec-98
SAMPLE.WB1 REV 2.0 22010NML			

4749

LABCORE Completed Worklist Report for Worklist# 27191

Analyst: scl Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: RERUN OF 26970,BY-105 szc

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@SR90-01 SR90-01 SOLID		1.04E-03	9.68E-4	93.077 % Recovery	
1 STD	0		@SR90-01 SR90-01C SOLID		100	8.62E+01	86.200 % Recovery	
1 STD	0		@SR90-01 SR90-01E SOLID		1.00	2.14E+02	2.140 % Ct. Error	
2 BLNK-PREP	0		@SR90-01 SR90-01 SOLID		1	2.73E-3	0.003 uCi/g	
2 BLNK-PREP	0		@SR90-01 SR90-01C SOLID		100	8.35E+01	83.500 % Recovery	
2 BLNK-PREP	0		@SR90-01 SR90-01E SOLID		1.00	4.13E+01	41.300 uCi/g	
3 BLNK/BKG	0		@SR90-01 SR90-01 SOLID		1	2.92E+00	2.920 BLNK/BKG	
4 SAMPLE	S98T002919	0 F	@SR90-01 SR90-01 SOLID		N/A	6.57E-01	1.14E-003 uCi/g	
4 SAMPLE	S98T002919	0 F	@SR90-01 SR90-01C SOLID		N/A	8.42E+01	% Recovery	
4 SAMPLE	S98T002919	0 F	@SR90-01 SR90-01E SOLID		N/A	1.85E+00	% Ct. Error	
5 DUP	S98T002919	0 F	@SR90-01 SR90-01 SOLID		6.57E-1	9.11E-1	32.398 RPD	
5 DUP	S98T002919	0 F	@SR90-01 SR90-01C SOLID		100	8.22E+01	82.200 % Recovery	
5 DUP	S98T002919	0 F	@SR90-01 SR90-01E SOLID		1.00	1.60E+00	1.600 % Cnt. Err	

Final page for worklist# 27191

Analyst Signature Date

Analyst Signature Date

19 Nov 98
Reviewer Signature Date

Results similar to last run, results accepted.

LABCORE Data Entry Template for Worklist# 27191

Analyst: S.L. Instrument: AB00 12 Book# 28057

Method: LA-220-101 Rev/Mod E - 3

Worklist Comment: RERUN OF 26970,BY-105 szc

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002919 0 F		@SR90-01	SOLID	98000490	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002919 0 F		@SR90-01	SOLID		

Final page for worklist # 27191

Sue Lee 11-18-98
Analyst Signature Date

new right 11/19/98
Analyst Signature Date
A. L. L. L. 11/10/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS				STANDARD	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000	
STD	TOTAL COUNTS (TC)	8456	GROSS WEIGHT (W2)	7.8851	
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.7989	
27191	BACKGROUND in cpm (BKG)	2.4	NET WEIGHT (W3)	0.0862	
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	7.68	
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	843.20	Sr-90 EFFICIENCY FACTO (C1)	0.4180	
Batch Number	CRITICAL LEVEL (Lc)	0.93	Y-90 EFFICIENCY FACTOR (C2)	0.4660	
98005806	TIME OF SEPARATION (ST)	10:42	Rmax	N/A	
ReRun	DATE OF SEPARATION (SD)	11/18/98	DETECTION LIMIT (Ld)	1.96	
0	TIME OF COUNT (TOC)	18:23	Sr-89/90 CONC. in µCi/L	9.6820E-01	
Sample Prep	DATE OF COUNT (DOC)	11/18/98			
N/A	STANDARD BOOK #	28B57			
Sample	STANDARD VALUE in µCi/mL	1.0427E-03			
WL27191-STD					
Instrument Code					
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
Prepared By	Sr-89/90 CONC in µCi/mL REPLACE RS WITH RMAX IF RS<=Lc AND RS>=0 OR REPLACE RS WITH Lc IF RS<0				
NEW	RS*DF*DDF*1000/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*REC*2220000				
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
SAC	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT)*1.96)				
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
SCL	NOTE: Expected weight = CVA * 0.1				
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
11/19/98	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
Analysis Date					
11/18/98	Sr-89/90 CONCENTRATION	9.68E-04	µCi/mL	DETECTION LEVEL 2.25E-06 µCi/L	
Analysis Time					
03:05 PM	RELATIVE COUNTING ERROR	2.1%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	86.2%			

Analyst:	SCL	Date:	19-Nov-98
Signature of Chemist:	SAC	Date:	

STANDARD.WB1 REV 2.0

22010NML

4752

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS (TC)	70	GROSS WEIGHT	(W2)	7.8770
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.7935
27191	BACKGROUND in cpm (BKG)	2.4	NET WEIGHT	(W3)	0.0835
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	7.87
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST FACTOR (g/L) (D g/L)	1.996			
SOLID	SAMPLE COUNT RATE (Rs)	4.60	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005806	TIME OF SEPARATION (ST)	10:42	Rmax		N/A
Run	DATE OF SEPARATION (SD)	11/18/98	DETECTION LIMIT	(Ld)	1.96
0	TIME OF COUNT (TOC)	18:34	Sr-89/90 CONC in $\mu\text{Ci/g}$		2.7268E-03
Sample Prep	DATE OF COUNT (DOC)	11/18/98			
N/A					
Sample #					
WL27191-BLK					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*(Dg/L)*REC*2220000)				
NEW	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/19/98					
Analysis Date					DETECTION LEVEL 1.16E-03 $\mu\text{Ci/g}$
11/18/98	Sr-89/90 CONCENTRATION	2.73E-03	$\mu\text{Ci/g}$		
Analysis Time					
03:05 PM	RELATIVE COUNTING ERROR	41.3%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	83.5%			

Analyst:	SCL	Date:	19-Nov-98
Signature of Chemist:	SAC	Date:	

BLANK.WB1 REV 2.0

22010NML

4753

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	11306	GROSS WEIGHT	(W2) 7.8743
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7901
27191	BACKGROUND in cpm	(BKG)	2.4	NET WEIGHT	(W3) 0.0842
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 8.75
@SR90-01	DILUTION FACTOR	DF	1		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.996		
SOLID	SAMPLE COUNT RATE	(Rs)	1128.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005806	TIME OF SEPARATION	(ST)	10:42	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	11/18/98	DETECTION LIMIT	(Ld) 1.96
1	TIME OF COUNT	(TOC)	19:27	Sr-89/90 CONC in $\mu\text{Ci/g}$	6.5735E-01
Sample Prep	DATE OF COUNT	(DOC)	11/18/98		
FUSION01					
Sample #					
S98T002919					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS < Lc and RS >= 0 or Replace RS with Lc if RS < 0				
Prepared By	RS * 1000 * DF / ((C1 + C2 * (1 - e to the power of ((-natural log 2) / (64.2 * DT)))) * SS * (Dg/L) * REC * 2220000)				
NEW	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/19/98					
Analysis Date					DETECTION LEVEL
11/18/98	Sr-89/90 CONCENTRATION	6.57E-01	$\mu\text{Ci/g}$		
Analysis Time					
03:05 PM	RELATIVE COUNTING ERROR	1.8%		1.14E-03	
Sample Point					$\mu\text{Ci/g}$
BY-105 (2)	PERCENT CARRIER RECOVERY	84.2%			

Analyst:	SCL	Date:	19-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

22010NML

4754

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	15107	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7991
27191	BACKGROUND in cpm	(BKG)	2.4	NET WEIGHT	(W3)	0.0822
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	8.92
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.97			
SOLID	SAMPLE COUNT RATE	(Rs)	1508.30	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005806	TIME OF SEPARATION	(ST)	10:42	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/18/98	DETECTION LIMIT	(Ld)	1.96
1	TIME OF COUNT	(TOC)	19:37	Sr-89/90 CONC in $\mu\text{Ci/g}$		9.1057E-01
Sample Prep	DATE OF COUNT	(DOC)	11/18/98			
FUSION01						
Sample #						
S98T002919						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*(Dg/L)*REC*2220000)					
NEW	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/19/98						
Analysis Date						DETECTION LEVEL 1.18E-03 $\mu\text{Ci/g}$
11/18/98	Sr-89/90 CONCENTRATION	9.11E-01		$\mu\text{Ci/g}$		
Analysis Time						
03:05 PM	RELATIVE COUNTING ERROR	1.6%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	82.2%				

Analyst:	SCL	Date:	19-Nov-98
Signature of Chemist:	SAC	Date:	

SAMPLE.WB1 REV 2.0

2201NML

4755

LABCORE Completed Worklist Report for Worklist# 27304

Analyst: gll Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: Rerun of 26840, BY-105 slh2

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		SR90-01 SR90-01 LIQUID		1.04E-03	1.06E-3	101.923 % Recovery	
1 STD	0		SR90-01 SR90-01C LIQUID		100	8.22E+01	82.200 % Recovery	
1 STD	0		SR90-01 SR90-01E LIQUID		1.00	2.09E+00	2.090 % Ct. Err	
2 BLNK	0		SR90-01 SR90-01 LIQUID		1	<2.27E-4	uCi/mL	
2 BLNK	0		SR90-01 SR90-01C LIQUID		100	8.58E+01	85.800 % Recovery	
2 BLNK	0		SR90-01 SR90-01E LIQUID		1.00	2.05E+02	205.000 uCi/mL	
3 BLNK/BKG	0		SR90-01 SR90-01 LIQUID		1	1.26E+00	1.260 BLNK/BKG	
4 SAMPLE	S98T003006	0	SR90-01 SR90-01 LIQUID		N/A	1.17E-01	2.56e-004 uCi/mL	
4 SAMPLE	S98T003006	0	SR90-01 SR90-01C LIQUID		N/A	8.58E+01	% Recovery	
4 SAMPLE	S98T003006	0	SR90-01 SR90-01E LIQUID		N/A	1.96E+00	% Ct. Error	
5 DUP	S98T003006	0	SR90-01 SR90-01 LIQUID		1.17E-1	1.21E-1	3.361 RPD	
5 DUP	S98T003006	0	SR90-01 SR90-01C LIQUID		100	8.67E+01	86.700 % Recovery	
5 DUP	S98T003006	0	SR90-01 SR90-01E LIQUID		1.00	1.91E+00	1.910 % Cnt Err	
6 SAMPLE	S98T003007	0	SR90-01 SR90-01 LIQUID		N/A	1.13E-01	2.49e-004 uCi/mL	
6 SAMPLE	S98T003007	0	SR90-01 SR90-01C LIQUID		N/A	8.79E+01	% Recovery	
6 SAMPLE	S98T003007	0	SR90-01 SR90-01E LIQUID		N/A	1.97E+00	% Ct. Error	
7 DUP	S98T003007	0	SR90-01 SR90-01 LIQUID		1.13E-1	1.15E-1	4.329 RPD	
7 DUP	S98T003007	0	SR90-01 SR90-01C LIQUID		100	8.49E+01	84.900 % Recovery	
7 DUP	S98T003007	0	SR90-01 SR90-01E LIQUID		1.00	1.95E+00	1.950 % Cnt Err	
8 SAMPLE	S98T003008	0	SR90-01 SR90-01 LIQUID		N/A	1.12E-01	2.56e-004 uCi/mL	
8 SAMPLE	S98T003008	0	SR90-01 SR90-01C LIQUID		N/A	8.48E+01	% Recovery	
8 SAMPLE	S98T003008	0	SR90-01 SR90-01E LIQUID		N/A	2.00E+00	% Ct. Error	
9 DUP	S98T003008	0	SR90-01 SR90-01 LIQUID		1.12E-1	1.13E-1	0.893 RPD	
9 DUP	S98T003008	0	SR90-01 SR90-01C LIQUID		100	8.54E+01	85.400 % Recovery	
9 DUP	S98T003008	0	SR90-01 SR90-01E LIQUID		1.00	1.98E+00	1.980 % Cnt Err	
10 SAMPLE	S98T003009	0	SR90-01 SR90-01 LIQUID		N/A	1.43E-01	2.53e-004 uCi/mL	
10 SAMPLE	S98T003009	0	SR90-01 SR90-01C LIQUID		N/A	8.56E+01	% Recovery	
10 SAMPLE	S98T003009	0	SR90-01 SR90-01E LIQUID		N/A	1.76E+00	% Ct. Error	
11 DUP	S98T003009	0	SR90-01 SR90-01 LIQUID		1.43E-1	1.41E-1	1.408 RPD	
11 DUP	S98T003009	0	SR90-01 SR90-01C LIQUID		100	8.37E+01	83.700 % Recovery	
11 DUP	S98T003009	0	SR90-01 SR90-01E LIQUID		1.00	1.79E+00	1.790 % Cnt Err	

Final page for worklist# 27304

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

 Reviewer Signature _____ Date 30 Nov 98

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27304

Analyst: gll Instrument: NONE _____ Book# 28057

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: Rerun of 26840, BY-105 slh2

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	LIQUID		
2 BLNK			@SR90-01	LIQUID		
3 BLNK/BKG			@SR90-01	LIQUID		
4 SAMPLE	S98T003006 0		@SR90-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T003006 0		@SR90-01	LIQUID		
6 SAMPLE	S98T003007 0		@SR90-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
7 DUP	S98T003007 0		@SR90-01	LIQUID		
8 SAMPLE	S98T003008 0		@SR90-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
9 DUP	S98T003008 0		@SR90-01	LIQUID		
10 SAMPLE	S98T003009 0		@SR90-01	LIQUID	98000510	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
11 DUP	S98T003009 0		@SR90-01	LIQUID		

Final page for worklist # 27304

gll 11-25-98
Analyst Signature Date

Shane L. Volden 11-26-98
Analyst Signature Date
MLB 11/26/98

Data Entry Comments:

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	8853	GROSS WEIGHT (W2)	7.8647
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.7825
27304	BACKGROUND in cpm (BKG)	3.1	NET WEIGHT (W3)	0.0822
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	7.92
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR (DDF)	1		
LIQUID	SAMPLE COUNT RATE (Rs)	882.20	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	1.06	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98005920	TIME OF SEPARATION (ST)	13:05	Rmax	N/A
Rerun	DATE OF SEPARATION (SD)	11/25/98	DETECTION LIMIT (Ld)	2.21
0	TIME OF COUNT (TOC)	21:00	Sr-89/90 CONC. in µCi/L 1.0598E+00	
Sample Prep	DATE OF COUNT (DOC)	11/25/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in µCi/mL	1.0422E-03		
WL27304-STD				
Instrument Code				
WB27811				
Prepared By	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
SLH2	Sr-89/90 CONC in µCi/mL REPLACE RS WITH RMAX IF RS<Lc AND RS>0 OR REPLACE RS WITH Lc IF RS<0			
Chemist	RS*DF*DDF*1000/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)			
SAC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))			
Analyst	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT)*1.96)			
GLL	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
Date Complete	NOTE: Expected weight = CVA * 0.1			
11/26/98	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
Analysis Date	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100			
11/25/98	Sr-89/90 CONCENTRATION	1.06E-03	µCi/mL	DETECTION LEVEL 2.66E-06 µCi/L
Analysis Time				
03:00 PM	RELATIVE COUNTING ERROR	2.1%		
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY	82.2%		

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	<i>SLH2</i>	Date:	30 Nov 98

STANDARD.WB1 REV 2.0

22010NML

4758

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	BLNK
BLNK	TOTAL COUNTS (TC)	39	GROSS WEIGHT	(W2)	7.8645
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.7787
27304	BACKGROUND in cpm (BKG)	3.1	NET WEIGHT	(W3)	0.0858
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	8.08
@SR90-01	DILUTION FACTOR (DF)	101			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	0.80	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION (ST)	13:05	Rmax		1.96
Rerun	DATE OF SEPARATION (SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT (TOC)	21:10	Sr-89/90 CONC in $\mu\text{Ci/L}$	<	2.2732E-01
Sample Prep	DATE OF COUNT (DOC)	11/25/98			
N/A					
Sample #					
WL27304-BLNK					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * DF * DDF / ((C1 + C2 * (1 - e^{-(\ln(2) * DT))}) * SS * REC * 2220000)$				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
11/26/98					
Analysis Date	Sr-89/90 CONCENTRATION < 2.27E-04 $\mu\text{Ci/mL}$				DETECTION LEVEL 2.57E-04 $\mu\text{Ci/mL}$
11/25/98	LESS THAN Value was Determined from Rmax.				
Analysis Time	RELATIVE COUNTING ERROR 205.0%				
03:00 PM					
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY 85.8%				

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98

BLANK.WB1 REV 2.0

22010NML

4759

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	10132	GROSS WEIGHT	(W2) 7.8653
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7795
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3) 0.0858
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 8.42
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1010.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005920	TIME OF SEPARATION	(ST)	13:05	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld) 2.21
0	TIME OF COUNT	(TOC)	21:30	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.1682E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98		
N/A					
Sample #					
S98T003006					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * DF * DDF / ((C1 + C2 * (1 - e \text{ to the power of } ((-\text{natural log } 2) / (64.2 * DT)))) * SS * REC * 2220000)$				
SLH2	NOTE: $64.2 = \text{Half Life for Y-90 and Rec.} = \text{Fractional Carrier Recovery } ((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $CVA * 0.1$				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
11/26/98					
Analysis Date					DETECTION LEVEL 2.56E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION	1.17E-01	$\mu\text{Ci/mL}$		
Analysis Time					
03:00 PM	RELATIVE COUNTING ERROR	2.0%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	85.8%			

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	10625	GROSS WEIGHT	(W2)	7.8945
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.8078
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0867
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	8.58
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1059.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:05	Rmax		N/A
Refun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	21:40	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.2105E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003006						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/26/98						
Analysis Date						DETECTION LEVEL 2.53E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION		1.21E-01	$\mu\text{Ci/mL}$		
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR		1.9%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		86.7%			

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0	22010NML		

4761

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	SAMPLE
SAMPLE	TOTAL COUNTS	(TC)	10039	GROSS WEIGHT	(W2)	7.8855
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7976
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0879
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	8.67
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1000.80	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:10	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	21:50	Sr-89/90 CONC in $\mu\text{Ci/L}$ 1.1270E+02		
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003007						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS <= Lc and RS >= 0 or Replace RS with Lc if RS < 0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/26/98						
Analysis Date						
11/25/98	Sr-89/90 CONCENTRATION	1.13E-01		$\mu\text{Ci/mL}$	DETECTION LEVEL	
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	2.0%		2.49E-04		
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	87.9%		$\mu\text{Ci/mL}$		

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4762

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	10173	GROSS WEIGHT	(W2)	7.8835
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7986
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0849
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.25
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1014.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:10	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	22:25	Sr-89/90 CONC in µCi/L		1.1756E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003007						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/26/98						
Analysis Date						DETECTION LEVEL 2.56E-04 µCi/mL
11/25/98	Sr-89/90 CONCENTRATION	1.18E-01		µCi/mL		
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	2.0%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	84.9%				

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4763

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	SAMPLE
SAMPLE	TOTAL COUNTS	(TC)	9655	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7767
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0848
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.42
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	962.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:10	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	22:35	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.1151E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003008						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/26/98						
Analysis Date						DETECTION LEVEL 2.56E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION	1.12E-01	$\mu\text{Ci/mL}$			
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	2.0%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	84.8%				

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4764

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	9902	GROSS WEIGHT	(W2) 7.8758
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.7904
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3) 0.0854
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 9.58
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	987.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98005920	TIME OF SEPARATION	(ST)	13:10	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld) 2.21
0	TIME OF COUNT	(TOC)	22:45	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.1338E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98		
N/A					
Sample #					
S98T003008					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)				
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
11/26/98					
Analysis Date					DETECTION LEVEL 2.54E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION	1.13E-01	$\mu\text{Ci/mL}$		
Analysis Time					
03:00 PM	RELATIVE COUNTING ERROR	2.0%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	85.4%			

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4765

WORKBOOK PAGE: SAM10

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	12552	GROSS WEIGHT	(W2)	7.8814
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7958
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0856
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.75
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1252.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:15	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	23:00	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.4325E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003009						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$					
Prepared By	$\text{RS} * \text{DF} * \text{DDF} / ((\text{C1} + \text{C2} * (1 - e^{-(\ln(2) * \text{DT})})) * \text{SS} * \text{REC} * 2220000)$					
SLH2	NOTE: $64.2 = \text{Half Life for Y-90 and Rec.} = \text{Fractional Carrier Recovery} ((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$					
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$					
11/26/98						
Analysis Date						DETECTION LEVEL 2.53E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION		1.43E-01	$\mu\text{Ci/mL}$		
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR		1.8%			
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		85.6%			

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30 Nov 98
SAMPLE.WB1 REV 2.0		22010NML	

4766

WORKBOOK PAGE: DUP11

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	12120	GROSS WEIGHT	(W2)	7.8787
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.7950
27304	BACKGROUND in cpm	(BKG)	3.1	NET WEIGHT	(W3)	0.0837
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.92
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1208.90	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.06	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98005920	TIME OF SEPARATION	(ST)	13:15	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	11/25/98	DETECTION LIMIT	(Ld)	2.21
0	TIME OF COUNT	(TOC)	23:10	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.4122E+02
Sample Prep	DATE OF COUNT	(DOC)	11/25/98			
N/A						
Sample #						
S98T003009						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
SLH2	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
11/26/98						
Analysis Date						DETECTION LEVEL 2.58E-04 $\mu\text{Ci/mL}$
11/25/98	Sr-89/90 CONCENTRATION	1.41E-01	$\mu\text{Ci/mL}$			
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	1.8%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	83.7%				

Analyst:	GLL	Date:	26-Nov-98
Signature of Chemist:	SAC	Date:	30/Nov 98
SAMPLE.WB1 REV 2.0		2201NOML	

4767

LABCORE Completed Worklist Report for Worklist# 27411

Analyst: scl Instrument: AB12 Book#

Method: Rev/Mod

Worklist Comment: Rerun of 26724 BY-105 szc

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 LIQUID		1.04E-03	1.03E-3	99.03% Recovery	
1 STD	0	@SR90-01 SR90-01C LIQUID		100	9.28E+01	92.800 % Recovery	
1 STD	0	@SR90-01 SR90-01E LIQUID		1.00	2.02E+00	2.020 % Ct. Err	
2 BLNK	0	@SR90-01 SR90-01 LIQUID		1	<2.01E-4	uCi/mL	
2 BLNK	0	@SR90-01 SR90-01C LIQUID		100	9.21E+01	92.100 % Recovery	
2 BLNK	0	@SR90-01 SR90-01E LIQUID		1.00	1.77E+02	177.000 uCi/mL	
3 BLNK/BKG	0	@SR90-01 SR90-01 LIQUID		1	1.36E+00	1.360 BLNK/BKG	
4 SAMPLE	S98T002814 0	@SR90-01 SR90-01 LIQUID		N/A	1.30E-01	2.07E-004 uCi/mL	
4 SAMPLE	S98T002814 0	@SR90-01 SR90-01C LIQUID		N/A	9.26E+01	% Recovery	
4 SAMPLE	S98T002814 0	@SR90-01 SR90-01E LIQUID		N/A	1.80E+00	% Ct. Error	
5 DUP	S98T002814 0	@SR90-01 SR90-01 LIQUID		1.30E-1	1.30E-1	0.000 RPD	
5 DUP	S98T002814 0	@SR90-01 SR90-01C LIQUID		100	9.29E+01	92.900 % Recovery	
5 DUP	S98T002814 0	@SR90-01 SR90-01E LIQUID		1.00	1.80E+00	1.800 % Ct. Err	
6 SAMPLE	S98T002815 0	@SR90-01 SR90-01 LIQUID		N/A	1.44E-01	2.06E-004 uCi/mL	
6 SAMPLE	S98T002815 0	@SR90-01 SR90-01C LIQUID		N/A	9.26E+01	% Recovery	
6 SAMPLE	S98T002815 0	@SR90-01 SR90-01E LIQUID		N/A	1.71E+00	% Ct. Error	
7 DUP	S98T002815 0	@SR90-01 SR90-01 LIQUID		1.44E-1	1.48E-1	2.740 RPD	
7 DUP	S98T002815 0	@SR90-01 SR90-01C LIQUID		100	9.25E+01	92.500 % Recovery	
7 DUP	S98T002815 0	@SR90-01 SR90-01E LIQUID		1.00	1.69E+00	1.690 % Ct. Err	
8 SAMPLE	S98T002816 0	@SR90-01 SR90-01 LIQUID		N/A	1.65E-01	2.07E-004 uCi/mL	
8 SAMPLE	S98T002816 0	@SR90-01 SR90-01C LIQUID		N/A	9.19E+01	% Recovery	
8 SAMPLE	S98T002816 0	@SR90-01 SR90-01E LIQUID		N/A	1.60E+00	% Ct. Error	
9 DUP	S98T002816 0	@SR90-01 SR90-01 LIQUID		1.65E-1	1.60E-1	3.077 RPD	
9 DUP	S98T002816 0	@SR90-01 SR90-01C LIQUID		100	9.05E+01	90.500 % Recovery	
9 DUP	S98T002816 0	@SR90-01 SR90-01E LIQUID		1.00	1.64E+00	1.640 % Ct. Err	
10 SAMPLE	S98T002817 0	@SR90-01 SR90-01 LIQUID		N/A	1.57E-01	2.04E-004 uCi/mL	
10 SAMPLE	S98T002817 0	@SR90-01 SR90-01C LIQUID		N/A	9.29E+01	% Recovery	
10 SAMPLE	S98T002817 0	@SR90-01 SR90-01E LIQUID		N/A	1.63E+00	% Ct. Error	
11 DUP	S98T002817 0	@SR90-01 SR90-01 LIQUID		1.57E-1	1.58E-1	0.635 RPD	
11 DUP	S98T002817 0	@SR90-01 SR90-01C LIQUID		100	9.33E+01	93.300 % Recovery	
11 DUP	S98T002817 0	@SR90-01 SR90-01E LIQUID		1.00	1.62E+00	1.620 % Ct. Err	

Final page for worklist# 27411

Analyst Signature Date
9 Dec 98
 Reviewer Signature Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27411

Analyst: S. L Instrument: AB00 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: Rerun of 26724 BY-105 szc

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	LIQUID		
2 BLNK			@SR90-01	LIQUID		
3 BLNK/BKG			@SR90-01	LIQUID		
4 SAMPLE	S98T002814 0		@SR90-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
5 DUP	S98T002814 0		@SR90-01	LIQUID		
6 SAMPLE	S98T002815 0		@SR90-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
7 DUP	S98T002815 0		@SR90-01	LIQUID		
8 SAMPLE	S98T002816 0		@SR90-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
9 DUP	S98T002816 0		@SR90-01	LIQUID		
10 SAMPLE	S98T002817 0		@SR90-01	LIQUID	98000437	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
11 DUP	S98T002817 0		@SR90-01	LIQUID		

Final page for worklist # 27411

Sue La 12-8-98
Analyst Signature Date

[Signature] 12/8/98
Analyst Signature Date
Sybil Z Chom 12/9/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	9455	GROSS WEIGHT (W2)	7.4304
WorkList	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3376
27411	BACKGROUND in cpm (BKG)	2.2	NET WEIGHT (W3)	0.0928
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	5.37
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR DDF	1		
LIQUID	SAMPLE COUNT RATE (Rs)	943.30	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.89	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98006034	TIME OF SEPARATION (ST)	08:48	Rmax	N/A
Rerun	DATE OF SEPARATION (SD)	12/08/98	DETECTION LIMIT (Ld)	1.88
0	TIME OF COUNT (TOC)	14:10	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0307E+00
Sample Prep	DATE OF COUNT (DOC)	12/08/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0414E-03		
WL27411-STD				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS \geq 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
MCB	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e^{-(\ln(2) * DT) / 64.2})) * SS * REC * 2220000)$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
SCL	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
12/09/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date				DETECTION LEVEL 2.05E-06 $\mu\text{Ci/L}$
12/08/98	Sr-89/90 CONCENTRATION	1.03E-03	$\mu\text{Ci/mL}$	
Analysis Time				
01:20 PM	RELATIVE COUNTING ERROR	2.0%		
Sample Point				
BY-105	PERCENT CARRIER RECOVERY	92.8%		

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98

STANDARD.WB1 REV 2.0

22010NML

4770

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)					BLNK
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
BLNK	TOTAL COUNTS	(TC)	30	GROSS WEIGHT	(W2) 7.4420
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3499
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0921
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 5.62
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	0.80	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	08:48	Rmax	1.81
ReRun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	14:25	Sr-89/90 CONC in $\mu\text{Ci/L}$	< 2.0053E-01
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
WL27411-BLNK					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					
12/08/98	Sr-89/90 CONCENTRATION	< 2.01E-04	$\mu\text{Ci/mL}$	DETECTION LEVEL	
Analysis Time	LESS THAN Value was Determined from Rmax.				
01:20 PM	RELATIVE COUNTING ERROR	176.7%	2.08E-04 $\mu\text{Ci/mL}$		
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	92.1%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98

BLANK.WB1 REV 2.0

22010NML

4771

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	11868	GROSS WEIGHT	(W2) 7.3928
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3002
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0926
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 5.78
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1184.60	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	08:48	Rmax	N/A
Retain	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	14:35	Sr-89/90 CONC in µCi/L	1.3043E+02
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
S98T002814					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					
12/08/98	Sr-89/90 CONCENTRATION	1.30E-01	µCi/mL	DETECTION LEVEL 2.07E-04 µCi/mL	
Analysis Time					
01:20 PM	RELATIVE COUNTING ERROR	1.8%			
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	92.6%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98

SAMPLE.WB1 REV 2.0   22010NML

4772

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	11878	GROSS WEIGHT	(W2) 7.4234
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3305
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0929
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 5.95
@SR90-01	DILUTION FACTOR	DF	101		
	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1185.60	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	08:48	Rmax	N/A
Retun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	14:45	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.2989E+02
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
S98T002814					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SLC	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					DETECTION LEVEL 2.06E-04 $\mu\text{Ci/mL}$
12/08/98	Sr-89/90 CONCENTRATION	1.30E-01	$\mu\text{Ci/mL}$		
Analysis Time					
01:20 PM	RELATIVE COUNTING ERROR	1.8%			
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	92.9%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0 2201NML			

4773

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000	
SAMPLE	TOTAL COUNTS (TC)	13134	GROSS WEIGHT	(W2)	7.3822	
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.2896	
27411	BACKGROUND in cpm (BKG)	2.2	NET WEIGHT	(W3)	0.0926	
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	6.08	
@SR90-01	DILUTION FACTOR (DF)	101				
Main	DIGEST DILUTION FACTOR (DDF)	1				
LIQUID	SAMPLE COUNT RATE (Rs)	1311.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL (Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98006034	TIME OF SEPARATION (ST)	08:55	Rmax		N/A	
ReRun	DATE OF SEPARATION (SD)	12/08/98	DETECTION LIMIT	(Ld)	1.88	
0	TIME OF COUNT (TOC)	15:00	Sr-89/90 CONC in µCi/L		1.4392E+02	
Sample Prep	DATE OF COUNT (DOC)	12/08/98				
N/A						
Sample #						
S98T002815						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.98					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/09/98						
Analysis Date						
12/08/98	Sr-89/90 CONCENTRATION	1.44E-01	µCi/mL	DETECTION LEVEL 2.06E-04 µCi/mL		
Analysis Time						
01:20 PM	RELATIVE COUNTING ERROR	1.7%				
Sample Point						
BY-105	PERCENT CARRIER RECOVERY	92.6%				

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4774

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Instrument Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	13519	GROSS WEIGHT	(W2)	7.4108
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3183
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	0.0925
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	6.33
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1349.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006034	TIME OF SEPARATION	(ST)	08:55	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	15:15	Sr-89/90 CONC in µCi/L		1.4791E+02
Sample Prep	DATE OF COUNT	(DOC)	12/08/98			
N/A						
Sample #						
S98T002815						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/09/98						
Analysis Date						DETECTION LEVEL 2.06E-04 µCi/mL
12/08/98	Sr-89/90 CONCENTRATION	1.48E-01	µCi/mL			
Analysis Time						
01:20 PM	RELATIVE COUNTING ERROR	1.7%				
Sample Point						
BY-105	PERCENT CARRIER RECOVERY	92.5%				

Analyst:		SCL	Date:	09-Dec-98
Signature of Chemist:		SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0		22010NML		

4775

WORKBOOK PAGE: SAM8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	14993	GROSS WEIGHT	(W2) 7.4071
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3152
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0919
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 6.58
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1497.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	08:55	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	15:30	Sr-89/90 CONC in µCi/L 1.6471E+02	
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
S98T002816					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					DETECTION LEVEL 2.07E-04 µCi/mL
12/08/98	Sr-89/90 CONCENTRATION	1.65E-01	µCi/mL		
Analysis Time					
01:20 PM	RELATIVE COUNTING ERROR	1.6%			
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	91.9%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE:WB1 REV 2.0		22010NML	

4776

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	14345	GROSS WEIGHT	(W2) 7.4012
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3107
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0905
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 6.75
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1432.28	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.90	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	08:55	Rmax	N/A
Return	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.89
0	TIME OF COUNT	(TOC)	15:40	Sr-89/90 CONC in µCi/L	1.5974E+02
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
S98T002816					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					
12/08/98	Sr-89/90 CONCENTRATION	1.60E-01	µCi/mL	DETECTION LEVEL	
Analysis Time					
01:20 PM	RELATIVE COUNTING ERROR	1.6%	2.10E-04 µCi/mL		
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	90.5%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4777

WORKBOOK PAGE: SAM10

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	14504	GROSS WEIGHT	(W2) 7.4030
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3101
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3) 0.0929
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 6.83
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	1448.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006034	TIME OF SEPARATION	(ST)	09:00	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld) 1.88
0	TIME OF COUNT	(TOC)	15:50	Sr-89/90 CONC in µCi/L 1.5721E+02	
Sample Prep	DATE OF COUNT	(DOC)	12/08/98		
N/A					
Sample #					
S98T002817					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)				
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.98				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/09/98					
Analysis Date					
12/08/98	Sr-89/90 CONCENTRATION	1.57E-01	µCi/mL		DETECTION LEVEL 2.04E-04 µCi/mL
Analysis Time					
01:20 PM	RELATIVE COUNTING ERROR	1.6%			
Sample Point					
BY-105	PERCENT CARRIER RECOVERY	92.9%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4778

WORKBOOK PAGE: DUP11

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	14720	GROSS WEIGHT	(W2)	7.3954
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3021
27411	BACKGROUND in cpm	(BKG)	2.2	NET WEIGHT	(W3)	0.0933
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	7.17
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	1469.80	Sr-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.89	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006034	TIME OF SEPARATION	(ST)	09:00	Rmax		N/A
ReRun	DATE OF SEPARATION	(SD)	12/08/98	DETECTION LIMIT	(Ld)	1.88
0	TIME OF COUNT	(TOC)	16:10	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.5832E+02
Sample Prep	DATE OF COUNT	(DOC)	12/08/98			
N/A						
Sample #						
S98T002817						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	$RS * DDF * ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / (64.2 * DT)))) * SS * REC * 2220000)$					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$					
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = $CVA * 0.1$					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$					
12/09/98						
Analysis Date						
12/08/98	Sr-89/90 CONCENTRATION		1.58E-01	$\mu\text{Ci/mL}$	DETECTION LEVEL	
Analysis Time						
01:20 PM	RELATIVE COUNTING ERROR		1.6%	2.02E-04 $\mu\text{Ci/mL}$		
Sample Point						
BY-105	PERCENT CARRIER RECOVERY		93.3%			

Analyst:	SCL	Date:	09-Dec-98
Signature of Chemist:	SAC	Date:	9 Dec 98
SAMPLE.WB1 REV 2.0 22010NML			

4779

LABCORE Completed Worklist Report for Worklist# 27546

Analyst: scl

Instrument: AB12

Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: Rerun of WL 26726,BY-105 mcb

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 LIQUID		1.04E-03	1.05E-3	100.962 % Recovery	
1 STD	0	@SR90-01 SR90-01C LIQUID		100	8.55E+01	85.500 % Recovery	
1 STD	0	@SR90-01 SR90-01E LIQUID		1.00	2.08E+00	2.080 % Ct. Error	
2 BLNK-PREP	0	@SR90-01 SR90-01 LIQUID		1	2.77E-4	2.770e-004 uCi/mL	
2 BLNK-PREP	0	@SR90-01 SR90-01C LIQUID		100	8.78E+01	87.800 % Recovery	
2 BLNK-PREP	0	@SR90-01 SR90-01E LIQUID		1.00	7.12E+01	71.200 uCi/mL	
3 BLNK/BKG	0	@SR90-01 SR90-01 LIQUID		1	1.92E+00	1.920 BLNK/BKG	
4 SAMPLE	S98T002960 0	@SR90-01 SR90-01 LIQUID		N/A	1.02E-01	2.30e-004 uCi/mL	
4 SAMPLE	S98T002960 0	@SR90-01 SR90-01C LIQUID		N/A	8.94E+01	% Recovery	
4 SAMPLE	S98T002960 0	@SR90-01 SR90-01E LIQUID		N/A	2.07E+00	% Ct. Error	
5 DUP	S98T002960 0	@SR90-01 SR90-01 LIQUID		1.02E-1	1.03E-1	0.976 RPD	
5 DUP	S98T002960 0	@SR90-01 SR90-01C LIQUID		100	8.81E+01	88.100 % Recovery	
5 DUP	S98T002960 0	@SR90-01 SR90-01E LIQUID		1.00	2.07E+00	2.070 % Ct. Err	
6 SAMPLE	S98T002961 0	@SR90-01 SR90-01 LIQUID		N/A	7.47E-02	2.25e-004 uCi/mL	
6 SAMPLE	S98T002961 0	@SR90-01 SR90-01C LIQUID		N/A	9.08E+01	% Recovery	
6 SAMPLE	S98T002961 0	@SR90-01 SR90-01E LIQUID		N/A	2.40E+00	% Ct. Error	
7 DUP	S98T002961 0	@SR90-01 SR90-01 LIQUID		7.47E-2	7.77E-2	3.937 RPD	
7 DUP	S98T002961 0	@SR90-01 SR90-01C LIQUID		100	8.56E+01	85.600 % Recovery	
7 DUP	S98T002961 0	@SR90-01 SR90-01E LIQUID		1.00	2.42E+00	2.420 % Ct. Err	
8 SAMPLE	S98T002962 0	@SR90-01 SR90-01 LIQUID		N/A	8.89E-02	2.25e-004 uCi/mL	
8 SAMPLE	S98T002962 0	@SR90-01 SR90-01C LIQUID		N/A	9.07E+01	% Recovery	
8 SAMPLE	S98T002962 0	@SR90-01 SR90-01E LIQUID		N/A	2.19E+00	% Ct. Error	
9 DUP	S98T002962 0	@SR90-01 SR90-01 LIQUID		8.89E-2	9.06E-2	1.894 RPD	
9 DUP	S98T002962 0	@SR90-01 SR90-01C LIQUID		100	8.87E+01	88.700 % Recovery	
9 DUP	S98T002962 0	@SR90-01 SR90-01E LIQUID		1.00	2.19E+00	2.190 % Ct. Err	
10 SAMPLE	S98T002963 0	@SR90-01 SR90-01 LIQUID		N/A	9.36E-02	2.25e-004 uCi/mL	
10 SAMPLE	S98T002963 0	@SR90-01 SR90-01C LIQUID		N/A	9.03E+01	% Recovery	
10 SAMPLE	S98T002963 0	@SR90-01 SR90-01E LIQUID		N/A	2.14E+00	% Ct. Error	
11 DUP	S98T002963 0	@SR90-01 SR90-01 LIQUID		9.36E-2	9.41E-2	0.533 RPD	
11 DUP	S98T002963 0	@SR90-01 SR90-01C LIQUID		100	8.81E+01	88.100 % Recovery	
11 DUP	S98T002963 0	@SR90-01 SR90-01E LIQUID		1.00	2.16E+00	2.160 % Ct. Err	
12 SAMPLE	S98T002964 0	@SR90-01 SR90-01 LIQUID		N/A	8.00E-02	2.24e-004 uCi/mL	
12 SAMPLE	S98T002964 0	@SR90-01 SR90-01C LIQUID		N/A	9.05E+01	% Recovery	
12 SAMPLE	S98T002964 0	@SR90-01 SR90-01E LIQUID		N/A	2.31E+00	% Ct. Error	
13 DUP	S98T002964 0	@SR90-01 SR90-01 LIQUID		8.00E-2	8.12E-2	1.489 RPD	
13 DUP	S98T002964 0	@SR90-01 SR90-01C LIQUID		100	8.43E+01	84.300 % Recovery	
13 DUP	S98T002964 0	@SR90-01 SR90-01E LIQUID		1.00	2.37E+00	2.370 % Ct. Err	

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27546

Analyst: S.L. Instrument: AB00 12 Book# 2F857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: Rerun of WL 26726,BY-105 mcb

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	LIQUID		
2 BLNK-PREP			@SR90-01	LIQUID		
3 BLNK/BKG			@SR90-01	LIQUID		
4 SAMPLE	S98T002960 0		@SR90-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
5 DUP	S98T002960 0		@SR90-01	LIQUID		
6 SAMPLE	S98T002961 0		@SR90-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
7 DUP	S98T002961 0		@SR90-01	LIQUID		
8 SAMPLE	S98T002962 0		@SR90-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
9 DUP	S98T002962 0		@SR90-01	LIQUID		
10 SAMPLE	S98T002963 0		@SR90-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
11 DUP	S98T002963 0		@SR90-01	LIQUID		
12 SAMPLE	S98T002964 0		@SR90-01	LIQUID	98000509	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
13 DUP	S98T002964 0		@SR90-01	LIQUID		

Final page for worklist # 27546

Sue Lan 12-30-98
Analyst Signature Date

Erin Buch 12-31-98
Analyst Signature Date
Spencer Brown 12/31/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	8916	GROSS WEIGHT (W2)	7.4214
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3359
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT (W3)	0.0855
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	6.17
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR (DDF)	1		
LIQUID	SAMPLE COUNT RATE (Rs)	889.00	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98006175	TIME OF SEPARATION (ST)	08:30	Rmax	N/A
Re-run	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT (Ld)	2.03
0	TIME OF COUNT (TOC)	14:40	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0454E+00
Sample Prep	DATE OF COUNT (DOC)	12/30/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0399E-03		
WL27546-STD				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS \geq 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
EMB	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / 64.2 * DT))) * SS * REC * 2220000)$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
SCL	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
12/31/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date	Sr-89/90 CONCENTRATION			DETECTION LEVEL
12/30/98	1.05E-03			$\mu\text{Ci/mL}$
Analysis Time	RELATIVE COUNTING ERROR			2.39E-06
11:30 AM	2.1%			$\mu\text{Ci/L}$
Sample Point	PERCENT CARRIER RECOVERY			
BY-105 (2)	85.5%			

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99

STANDARD.WB1 REV 2.0

22010NML

4783

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	BLNK-PREP
BLNK-PREP	TOTAL COUNTS (TC)	50	GROSS WEIGHT	(W2)	7.3959
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3081
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT	(W3)	0.0878
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	6.33
@SR90-01	DILUTION FACTOR (DF)	101			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	2.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION (ST)	08:30	Rmax		N/A
ReRun	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT (TOC)	14:50	Sr-89/90 CONC in $\mu\text{Ci/L}$		2.7710E-01
Sample Prep	DATE OF COUNT (DOC)	12/30/98			
N/A					
Sample #					
WL27546 BLNK-PREP					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS < Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * DF * DDF / ((C1 + C2 * (1 - e^{-((-\ln(2) * DT))}) * SS * REC * 2220000))$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analys	NOTE: Expected weight = $CVA * 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
12/31/98					
Analysis Date					
12/30/98	Sr-89/90 CONCENTRATION	2.77E-04	$\mu\text{Ci/mL}$	DETECTION LEVEL	
Analysis Time					
11:30 AM	RELATIVE COUNTING ERROR	71.2%	2.35E-04		
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	87.8%	$\mu\text{Ci/mL}$		

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99

BLANK.WB1 REV 2.0

22010NML

4784

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	9041	GROSS WEIGHT	(W2)	7.3994
Worklist	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3100
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT	(W3)	0.0894
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	6.50
@SR90-01	DILUTION FACTOR (DF)	101			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	901.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION (ST)	08:30	Rmax		N/A
Refun	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT (TOC)	15:00	Sr-89/90 CONC in $\mu\text{Ci/L}$ 1.0204E+02		
Sample/Rep	DATE OF COUNT (DOC)	12/30/98			
N/A					
Sample #					
S98T002960					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS \geq 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS \cdot DF \cdot DDF / ((C1 + C2 \cdot (1 - e^{-(\ln(2) \cdot DT / 64.2)})) \cdot SS \cdot REC \cdot 2220000)$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG \cdot CT) / (TC - BKG \cdot CT)$) $\cdot 1.96$				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) $\cdot 100$				
Analyst	NOTE: Expected weight = $CVA \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) \cdot 24) + (TOC - ST) / 100$				
12/31/98					
Analysis Date	Sr-89/90 CONCENTRATION 1.02E-01 $\mu\text{Ci/mL}$				DETECTION LEVEL 2.30E-04 $\mu\text{Ci/mL}$
Analysis Time	RELATIVE COUNTING ERROR 2.1%				
11:30 AM					
Sample Point	PERCENT CARRIER RECOVERY 89.4%				
BY-105 (2)					

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99
SAMPLE.WB1 REV 2.0		22010NML	

4785

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS (TC)	9001	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT	(W1)	7.3691
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT	(W3)	7.2810
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS)	(DT)	0.0881
@SR90-01	DILUTION FACTOR (DF)	101			6.67
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	897.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION (ST)	08:30	Rmax		N/A
Retun	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT (TOC)	15:10	Sr-89/90 CONC in $\mu\text{Ci/L}$		1.0291E+02
Sample Prep	DATE OF COUNT (DOC)	12/30/98			
N/A					
Sample #					
S98T002960					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS \cdot DF \cdot DDF / ((C1 + C2 \cdot (1 - e^{-(\ln(2) / (64.2 \cdot DT))})) \cdot SS \cdot REC \cdot 2220000)$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG \cdot CT) / (TC - BKG \cdot CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $CVA \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) \cdot 24) + (TOC - ST) / 100$				
12/31/98					
Analysis Date					
12/30/98	Sr-89/90 CONCENTRATION	1.03E-01	$\mu\text{Ci/mL}$	DETECTION LEVEL	
Analysis Time					
11:30 AM	RELATIVE COUNTING ERROR	2.1%	2.33E-04 $\mu\text{Ci/mL}$		
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	88.1%			

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99
SAMPLE WB1 REV 2.0		22010NML	

4786

WORKBOOK PAGE: SAM6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	6772	GROSS WEIGHT	(W2)	7.4021
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3113
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	0.0908
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	7.08
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	674.60	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION	(ST)	08:35	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT	(TOC)	15:40	Sr-89/90 CONC in $\mu\text{Ci/L}$		7.4730E+01
Sample Prep	DATE OF COUNT	(DOC)	12/30/98			
N/A						
Sample #						
S98T002961						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/31/98						
Analysis Date	Sr-89/90 CONCENTRATION					DETECTION LEVEL 2.25E-04 $\mu\text{Ci/mL}$
12/30/98	7.47E-02 $\mu\text{Ci/mL}$					
Analysis Time						
11:30 AM	RELATIVE COUNTING ERROR					
Sample Point	2.4%					
BY-105 (2)	PERCENT CARRIER RECOVERY					
	90.8%					

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99
SAMPLE:WB1 REV 2.0 22010NML			

4787

WORKBOOK PAGE: DUP7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	6653	GROSS WEIGHT	(W2)	7.4052
Work Use	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3196
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	0.0856
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	7.25
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	662.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION	(ST)	08:35	Rmax		N/A
Refun	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT	(TOC)	15:50	Sr-89/90 CONC in $\mu\text{Ci/L}$ 7.7738E+01		
Sample Prep	DATE OF COUNT	(DOC)	12/30/98			
N/A						
Sample #						
S98T002961						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/31/98						
Analysis Date						DETECTION LEVEL
12/30/98	Sr-89/90 CONCENTRATION		7.77E-02		$\mu\text{Ci/mL}$	
Analysis Time						
11:30 AM	RELATIVE COUNTING ERROR		2.4%		2.39E-04	
Sample Point						$\mu\text{Ci/mL}$
BY-105 (2)	PERCENT CARRIER RECOVERY		85.6%			

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99
SAMPLE.WB1 REV 2.0		22010NML	

4788

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	8073	GROSS WEIGHT (W2)		7.4134
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.3227
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT (W3)		0.0907
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)		7.42
@SR90-01	DILUTION FACTOR (DF)	101			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	804.70	SR-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98006175	TIME OF SEPARATION (ST)	08:35	Rmax		N/A
Refun	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT (Ld)		2.03
0	TIME OF COUNT (TOC)	16:00	Sr-89/90 CONC in $\mu\text{Ci/L}$ 8.8935E+01		
Sample Rep	DATE OF COUNT (DOC)	12/30/98			
N/A					
Sample #					
S98T002962					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * DF * DDF / ((C1 + C2 * (1 - e^{-(\ln(2) * DT) / 64.2 * DT)})) * SS * REC * 2220000$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $CVA * 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
12/31/98					
Analysis Date	Sr-89/90 CONCENTRATION	8.89E-02	$\mu\text{Ci/mL}$	DETECTION LEVEL	
12/30/98					
Analysis Time	RELATIVE COUNTING ERROR	2.2%	2.25E-04 $\mu\text{Ci/mL}$		
11:30 AM					
Sample Point	PERCENT CARRIER RECOVERY	90.7%			
BY-105 (2)					

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99

SAMPLE.WB1 REV 2.0

22010NML

4789

WORKBOOK PAGE: DUP9

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP
DUP	TOTAL COUNTS	(TC)	8053	GROSS WEIGHT	(W2)	1.000
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.4079
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	7.3192
Task Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	0.0887
@SR90-01	DILUTION FACTOR	DF	101			7.58
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	802.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION	(ST)	08:35	Rmax		N/A
Refun	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT	(TOC)	16:10	Sr-89/90 CONC in $\mu\text{Ci/L}$		9.0560E+01
Sample Prep	DATE OF COUNT	(DOC)	12/30/98			
N/A						
Sample #						
S98T002962						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/31/98						
Analysis Date						
12/30/98	Sr-89/90 CONCENTRATION	9.06E-02		$\mu\text{Ci/mL}$		DETECTION LEVEL 2.29E-04 $\mu\text{Ci/mL}$
Analysis Time						
11:30 AM	RELATIVE COUNTING ERROR	2.2%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	88.7%				

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99
SAMPLE WB1 REV 2.0		22010NML	

4790

WORKBOOK PAGE: SAM10

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	8477	GROSS WEIGHT	(W2) 7.4163
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3260
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3) 0.0903
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 7.67
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	845.10	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006175	TIME OF SEPARATION	(ST)	08:40	Rmax	N/A
ReRun	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld) 2.03
0	TIME OF COUNT	(TOC)	16:20	Sr-89/90 CONC in $\mu\text{Ci/L}$	9.3575E+01
Sample Prep	DATE OF COUNT	(DOC)	12/30/98		
N/A					
Sample #					
S98T002963					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} \geq 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} \cdot \text{DF} \cdot \text{DDF} / ((\text{C1} + \text{C2} \cdot (1 - e^{(-(\text{natural log } 2) / (64.2 \cdot \text{DT})))) \cdot \text{SS} \cdot \text{REC} \cdot 2220000)$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) \cdot 24) + (\text{TOC} - \text{ST}) / 100$				
12/31/98					
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL
12/30/98	9.36E-02				$\mu\text{Ci/mL}$
Analysis Time	RELATIVE COUNTING ERROR				2.25E-04
11:30 AM	2.1%				$\mu\text{Ci/mL}$
Sample Point	PERCENT CARRIER RECOVERY				
BY-105 (2)	90.3%				

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99
SAMPLE.WB1 REV 2.0		22010NML	

4791

WORKBOOK PAGE: DUP11

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	8337	GROSS WEIGHT	(W2)	7.4054
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3173
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	0.0881
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	7.92
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	831.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006175	TIME OF SEPARATION	(ST)	08:40	Rmax		N/A
Return	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03
0	TIME OF COUNT	(TOC)	16:35	Sr-89/90 CONC in $\mu\text{Ci/L}$ 9.4083E+01		
Sample Prep	DATE OF COUNT	(DOC)	12/30/98			
N/A						
Sample #						
S98T002963						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/31/98						
Analysis Date						DETECTION LEVEL 2.30E-04 $\mu\text{Ci/mL}$
12/30/98	Sr-89/90 CONCENTRATION	9.41E-02		$\mu\text{Ci/mL}$		
Analysis Time						
11:30 AM	RELATIVE COUNTING ERROR	2.2%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	88.1%				

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99
SAMPLE.WB1 REV 2.0 22010NML			

4792

WORKBOOK PAGE: SAM12

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				SAMPLE	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS (TC)	7298	GROSS WEIGHT (W2)		7.4267
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.3362
27546	BACKGROUND in cpm (BKG)	2.6	NET WEIGHT (W3)		0.0905
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)		8.08
@SR90-01	DILUTION FACTOR (DF)	101			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	727.20	SR-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.97	Y-90 EFFICIENCY FACTOR (C2)		0.4660
98006175	TIME OF SEPARATION (ST)	08:40	Rmax		N/A
Rec'd	DATE OF SEPARATION (SD)	12/30/98	DETECTION LIMIT (Ld)		2.03
0	TIME OF COUNT (TOC)	16:45	Sr-89/90 CONC in $\mu\text{Ci/L}$		8.0004E+01
Sample Prep	DATE OF COUNT (DOC)	12/30/98			
N/A					
Sample #					
S98T002964					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} \cdot \text{DF} \cdot \text{DDF} / ((\text{C1} + \text{C2} \cdot (1 - e^{-(\ln(2) \cdot \text{DT})})) \cdot \text{SS} \cdot \text{REC} \cdot 2220000)$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} \cdot 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} \cdot \text{CT}) / (\text{TC} - \text{BKG} \cdot \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} \cdot 0.1$				
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) \cdot 24) + (\text{TOC} - \text{ST}) / 100$				
12/31/98					
Analysis Date					DETECTION LEVEL
12/30/98	Sr-89/90 CONCENTRATION	8.00E-02	$\mu\text{Ci/mL}$		
Analysis Time					
11:30 AM	RELATIVE COUNTING ERROR	2.3%			
Sample Point					2.24E-04 $\mu\text{Ci/mL}$
BY-105 (2)	PERCENT CARRIER RECOVERY	90.5%			

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	7 Jan 99
SAMPLE.WB1 REV 2.0	22010NML		

4793

WORKBOOK PAGE: DUP13

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	DUP	
DUP	TOTAL COUNTS	(TC)	6915	GROSS WEIGHT	(W2)	1.000	
Work Unit	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.4214	
27546	BACKGROUND in cpm	(BKG)	2.6	NET WEIGHT	(W3)	7.3371	
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	0.0843	
@SR90-01	DILUTION FACTOR	DF	101			8.33	
Matrix	DIGEST DILUTION FACTOR	(DDF)	1				
LIQUID	SAMPLE COUNT RATE	(Rs)	688.90	SR-90 EFFICIENCY FACTOR	(C1)	0.4180	
Batch Number	CRITICAL LEVEL	(Lc)	0.97	Y-90 EFFICIENCY FACTOR	(C2)	0.4660	
98006175	TIME OF SEPARATION	(ST)	08:40	Rmax		N/A	
Refun	DATE OF SEPARATION	(SD)	12/30/98	DETECTION LIMIT	(Ld)	2.03	
0	TIME OF COUNT	(TOC)	17:00	Sr-89/90 CONC in µCi/L		8.1160E+01	
Sample Prep	DATE OF COUNT	(DOC)	12/30/98				
N/A							
Sample							
S98T002964							
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)						
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS <= Lc and RS >= 0 or Replace RS with Lc if RS < 0						
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)						
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))						
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96						
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100						
Analyst	NOTE: Expected weight = CVA * 0.1						
SCL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.						
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100						
12/31/98							
Analysis Date	Sr-89/90 CONCENTRATION					DETECTION LEVEL	
12/30/98	8.12E-02						
Analysis Time							
11:30 AM	RELATIVE COUNTING ERROR						
Sample Point	2.4%					2.40E-04	
BY-105 (2)	PERCENT CARRIER RECOVERY					µCi/mL	
	84.3%						

Analyst:	SCL	Date:	31-Dec-98
Signature of Chemist:	SAC	Date:	4 Jan 99
SAMPLE.WB1 REV 2.0	22010NML		

4794

LABCORE Completed Worklist Report for Worklist# 27669

Analyst: gl Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY-105 FOR @SR90-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@SR90-01 SR90-01 LIQUID		1.04E-03	1.04E-3	100.000 % Recovery	
1 STD	0		@SR90-01 SR90-01C LIQUID		100	9.27E+01	92.700 % Recovery	
1 STD	0		@SR90-01 SR90-01E LIQUID		1.00	2.01E+00	2.010 % Ct. Err	Extr
2 BLNK	0		@SR90-01 SR90-01 LIQUID		1	<1.54E-3		uCi/mL
2 BLNK	0		@SR90-01 SR90-01C LIQUID		100	9.20E+01	92.000 % Recovery	
2 BLNK	0		@SR90-01 SR90-01E LIQUID		1.00	3.53E+02	353.000	uCi/mL
3 BLNK/BKG	0		@SR90-01 SR90-01 LIQUID		1	1.17E+00	1.170	BLNK/BKG
4 SAMPLE	S98T003384	0	@SR90-01 SR90-01 LIQUID		N/A	1.26E-01	2.16E-003	uCi/mL
4 SAMPLE	S98T003384	0	@SR90-01 SR90-01C LIQUID		N/A	9.20E+01		% Recovery
4 SAMPLE	S98T003384	0	@SR90-01 SR90-01E LIQUID		N/A	5.92E+00		% Ct. Error
5 DUP	S98T003384	0	@SR90-01 SR90-01 LIQUID		1.26E-1	1.32E-1	4.651	RPD
5 DUP	S98T003384	0	@SR90-01 SR90-01C LIQUID		100	9.15E+01	91.500 % Recovery	
5 DUP	S98T003384	0	@SR90-01 SR90-01E LIQUID		1.00	5.80E+00	5.800 % Ct. Err	

Final page for worklist# 27669

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

27 Dec 98
Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27669

Analyst: gll Instrument: AB00 12 Book# 28857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY-105 FOR @SR90-01 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@SR90-01	LIQUID		
2	BLNK				@SR90-01	LIQUID		
3	BLNK/BKG				@SR90-01	LIQUID		
4	SAMPLE	S98T003384 0			@SR90-01	LIQUID	98000437	BY-105 (2)
					Analytes Requested: SR90-01 , SR90-01C, SR90-01E			
5	DUP	S98T003384 0			@SR90-01	LIQUID		

Final page for worklist # 27669

[Signature] 12-23-98
Analyst Signature Date

[Signature] 12/28/98
Analyst Signature Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	9555	GROSS WEIGHT (W2)	7.3900
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.2973
27669	BACKGROUND in cpm (BKG)	2.4	NET WEIGHT (W3)	0.0927
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	5.58
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR DDF	1		
LIQUID	SAMPLE COUNT RATE (Rs)	953.10	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	0.93	Y-90 EFFICIENCY FACTOR (C2)	0.4660
98006306	TIME OF SEPARATION (ST)	08:20	Rmax	N/A
Regrin	DATE OF SEPARATION (SD)	12/23/98	DETECTION LIMIT (Ld)	1.96
0	TIME OF COUNT (TOC)	13:55	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0401E+00
Sample Prep	DATE OF COUNT (DOC)	12/23/98		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0403E-03		
WL27669				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS > 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
SZC	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / 64.2 * DT)))) * SS * REC * 2220000$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$			
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT) * 1.96)$			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
GLL	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
12/28/98	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$			
Analysis Date	Sr-89/90 CONCENTRATION	1.04E-03	$\mu\text{Ci/mL}$	DETECTION LEVEL
12/23/98				
Analysis Time	RELATIVE COUNTING ERROR	2.0%		2.14E-06
10:25 AM				$\mu\text{Ci/L}$
Sample Point	PERCENT CARRIER RECOVERY	92.7%		
BY-105 (2)				

Analyst:	GLL	Date:	28-Dec-98
Signature of Chemist:	SAC	Date:	29 Dec 98

STANDARD.WB1 REV 2.0

22010NML

4797

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK	TOTAL COUNTS	(TC)	28	GROSS WEIGHT	(W2)	7.3868
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.2948
27669	BACKGROUND in cpm	(BKG)	2.4	NET WEIGHT	(W3)	0.0920
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT)	6.00
@SR90-01	DILUTION FACTOR	DF	101			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	0.40	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006306	TIME OF SEPARATION	(ST)	08:20	Rmax		1.39
Rerun	DATE OF SEPARATION	(SD)	12/23/98	DETECTION LIMIT	(Ld)	1.96
0	TIME OF COUNT	(TOC)	14:20	Sr-89/90 CONC in $\mu\text{Ci/L}$	<	1.5370E+00
Sample Prep	DATE OF COUNT	(DOC)	12/23/98			
N/A						
Sample #						
WL27669						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT)))*SS*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery (W2-W1) / (CVA * 0.1000)					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
12/28/98						
Analysis Date						DETECTION LEVEL 2.16E-03 $\mu\text{Ci/mL}$
12/23/98	Sr-89/90 CONCENTRATION	< 1.54E-03	$\mu\text{Ci/mL}$			
Analysis Time	LESS THAN Value was Determined from Rmax.					
10:25 AM	RELATIVE COUNTING ERROR	353.3%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	92.0%				

Analyst:	GLL	Date:	28-Dec-98
Signature of Chemist:	SAC	Date:	29 Dec 98
BLANK.WB1 REV 2.0 22010NML			

4798

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC)	1166	GROSS WEIGHT	(W2) 7.4120
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3200
27669	BACKGROUND in cpm	(BKG)	2.4	NET WEIGHT	(W3) 0.0920
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 6.28
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	114.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006306	TIME OF SEPARATION	(ST)	08:20	Rmax	N/A
Recm	DATE OF SEPARATION	(SD)	12/23/98	DETECTION LIMIT	(Ld) 1.96
0	TIME OF COUNT	(TOC)	14:37	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.2590E+02
Sample Prep	DATE OF COUNT	(DOC)	12/23/98		
N/A					
Sample #					
S98T003384					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * DDF * DDF / ((C1 + C2 * (1 - e \text{ to the power of } ((-\text{natural log } 2) / (64.2 * DT)))) * SS * REC * 2220000)$				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/28/98					
Analysis Date					DETECTION LEVEL 2.16E-03 $\mu\text{Ci/mL}$
12/23/98	Sr-89/90 CONCENTRATION	1.26E-01	$\mu\text{Ci/mL}$		
Analysis Time					
10:25 AM	RELATIVE COUNTING ERROR	5.9%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	92.0%			

Analyst:	GLL	Date:	28-Dec-98
Signature of Chemist:	SAC	Date:	29 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4799

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	1213	GROSS WEIGHT	(W2) 7.4040
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3125
27669	BACKGROUND in cpm	(BKG)	2.4	NET WEIGHT	(W3) 0.0915
Test Code	SAMPLE VOLUME in mL	(SS)	0.100	DELTA TIME (HOURS)	(DT) 6.47
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST DILUTION FACTOR	(DDF)	1		
LIQUID	SAMPLE COUNT RATE	(Rs)	118.90	SR-90 EFFICIENCY FACTOR	(C1) 0.4180
Batch Number	CRITICAL LEVEL	(Lc)	0.93	Y-90 EFFICIENCY FACTOR	(C2) 0.4660
98006306	TIME OF SEPARATION	(ST)	08:20	Rmax	N/A
Rerun	DATE OF SEPARATION	(SD)	12/23/98	DETECTION LIMIT	(Ld) 1.96
0	TIME OF COUNT	(TOC)	14:48	Sr-89/90 CONC in $\mu\text{Ci/L}$	1.3154E+02
Sample Prep	DATE OF COUNT	(DOC)	12/23/98		
N/A					
Sample #					
S98T003384					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in $\mu\text{Ci/L}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
12/28/98					
Analysis Date					DETECTION LEVEL 2.17E-03 $\mu\text{Ci/mL}$
12/23/98	Sr-89/90 CONCENTRATION	1.32E-01	$\mu\text{Ci/mL}$		
Analysis Time					
10:25 AM	RELATIVE COUNTING ERROR	5.8%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	91.5%			

Analyst:	GLL	Date:	28-Dec-98
Signature of Chemist:	SAC	Date:	29 Dec 98
SAMPLE.WB1 REV 2.0		22010NML	

4800

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 27743

Analyst: crj Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: RERUN OF 27139 BY-105 szc

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@SR90-01 SR90-01 SOLID		1.04E-03	1.02E-3	98.077 % Recovery	
1 STD	0		@SR90-01 SR90-01C SOLID		100	9.33E+01	93.300 % Recovery	
1 STD	0		@SR90-01 SR90-01E SOLID		1.00	1.99E+00	1.990 % Ct. Error	
2 BLNK-PREP	0		@SR90-01 SR90-01 SOLID		1	<1.34E-3		uCi/g
2 BLNK-PREP	0		@SR90-01 SR90-01C SOLID		100	9.28E+01	92.800 % Recovery	
2 BLNK-PREP	0		@SR90-01 SR90-01E SOLID		1.00	3.61E+02	361.000	uCi/g
3 BLNK/BKG	0		@SR90-01 SR90-01 SOLID		1	8.89E-01	0.889	BLNK/BKG
4 SAMPLE	S98T002988	0 F	@SR90-01 SR90-01 SOLID		N/A	4.91E+00	2.76E-003	uCi/g
4 SAMPLE	S98T002988	0 F	@SR90-01 SR90-01C SOLID		N/A	9.32E+01		% Recovery
4 SAMPLE	S98T002988	0 F	@SR90-01 SR90-01E SOLID		N/A	9.05E-01		% Ct. Error
5 DUP	S98T002988	0 F	@SR90-01 SR90-01 SOLID		4.91E+00	3.50E+00	33.532	RPD
5 DUP	S98T002988	0 F	@SR90-01 SR90-01C SOLID		100	9.31E+01	93.100 % Recovery	
5 DUP	S98T002988	0 F	@SR90-01 SR90-01E SOLID		1.00	1.07E+00	1.070 % Ct. Error	

Final page for worklist# 27743

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

Reviewer Signature _____
5 Jan 99
Date _____

Result similar to 1st run. Result accepted. me

LABCORE Data Entry Template for Worklist# 27743

Analyst: C25 Instrument: AB00 12 Book# 28357

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: RERUN OF 27139 BY-105 szc

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 BLNK/BKG			@SR90-01	SOLID		
4 SAMPLE	S98T002988	0 F	@SR90-01	SOLID	98000509	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002988	0 F	@SR90-01	SOLID		

Final page for worklist # 27743

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD	
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000	
STD	TOTAL COUNTS (TC)	9825	GROSS WEIGHT (W2)	7.4057	
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3124	
27743	BACKGROUND in cpm (BKG)	4.5	NET WEIGHT (W3)	0.0933	
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	9.52	
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST DILUTION FACTOR DDF	1			
LIQUID	SAMPLE COUNT RATE (Rs)	978.00	SR-90 EFFICIENCY FACTO (C1)	0.4180	
Batch Number	CRITICAL LEVEL (Lc)	1.28	Y-90 EFFICIENCY FACTOR (C2)	0.4660	
98006379	TIME OF SEPARATION (ST)	10:54	Rmax	N/A	
Rerun	DATE OF SEPARATION (SD)	12/31/98	DETECTION LIMIT (Ld)	2.65	
0	TIME OF COUNT (TOC)	20:25	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0187E+00	
Sample Prep	DATE OF COUNT (DOC)	12/31/98			
N/A	STANDARD BOOK #	28B57			
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0398E-03			
WL27743-STD					
Instrument Code					
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF RS<Lc AND RS>=0 OR REPLACE RS WITH Lc IF RS<0				
MCB	RS*DF*DDF*1000/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)				
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
SAC	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT)*1.96)				
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
CRJ	NOTE: Expected weight = CVA * 0.1				
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
01/04/99	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL 2.76E-06 $\mu\text{Ci/L}$
12/31/98	1.02E-03 $\mu\text{Ci/mL}$				
Analysis Time	RELATIVE COUNTING ERROR				
01:35 PM	2.0%				
Sample Point					
BY-105	PERCENT CARRIER RECOVERY				93.3%

Analyst:	CRJ	Date:	04-Jan-99
Signature of Chemist:	SAC	Date:	5 Jan 99

STANDARD.WB1 REV 2.0

22010NML

4863

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	40	GROSS WEIGHT	(W2)	7.4454
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3526
27743	BACKGROUND in cpm	(BKG)	4.5	NET WEIGHT	(W3)	0.0928
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	9.77
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9872			
SOLID	SAMPLE COUNT RATE	(Rs)	-0.50	Sr-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.28	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006379	TIME OF SEPARATION	(ST)	10:54	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/31/98	DETECTION LIMIT	(Ld)	2.65
0	TIME OF COUNT	(TOC)	20:40	Sr-89/90 CONC in $\mu\text{Ci/g}$	<	1.3438E-03
Sample Prep	DATE OF COUNT	(DOC)	12/31/98			
N/A						
Sample #						
WL27743-BLNC						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} \geq 0$ or Replace RS with Lc if $\text{RS} < 0$					
Prepared By	$\text{RS} * 1000 * \text{DF} / ((\text{C1} + \text{C2} * (1 - e \text{ to the power of } ((-\text{natural log } 2) / (64.2 * \text{DT})))) * \text{SS} * (\text{Dg/L}) * \text{REC} * 2220000)$					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $(\text{W2} - \text{W1}) / (\text{CVA} * 0.1000)$					
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$					
01/04/99						
Analysis Date						
12/31/98	Sr-89/90 CONCENTRATION	< 1.34E-03				DETECTION LEVEL 2.78E-03 $\mu\text{Ci/g}$
Analysis Time	LESS THAN Value was Determined from Lc.					
01:35 PM	RELATIVE COUNTING ERROR	361.4%				
Sample Point						
BY-105	PERCENT CARRIER RECOVERY	92.8%				

Analyst:	CRJ	Date:	04-Jan-99
Signature of Chemist:	SAC	Date:	5 Jan 99

BLANK.WB1 REV 2.0

22010NML

4804

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

						SAMPLE
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	47088	GROSS WEIGHT	(W2)	7.4229
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3297
27743	BACKGROUND in cpm	(BKG)	4.5	NET WEIGHT	(W3)	0.0932
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	10.10
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9872			
SOLID	SAMPLE COUNT RATE	(Rs)	4704.30	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.28	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006379	TIME OF SEPARATION	(ST)	10:54	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/31/98	DETECTION LIMIT	(Ld)	2.65
1	TIME OF COUNT	(TOC)	21:00	Sr-89/90 CONC in $\mu\text{Ci/g}$		4.9090E+00
Sample Prep	DATE OF COUNT	(DOC)	12/31/98			
FUSION01						
Sample #						
S98T002988						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*22200000					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
01/04/99						
Analysis Date						DETECTION LEVEL 2.76E-03 $\mu\text{Ci/g}$
12/31/98	Sr-89/90 CONCENTRATION	4.91E+00	$\mu\text{Ci/g}$			
Analysis Time						
01:35 PM	RELATIVE COUNTING ERROR	0.9%				
Sample Point						
BY-105	PERCENT CARRIER RECOVERY	93.2%				

Analyst:	CRJ	Date:	04-Jan-99
Signature of Chemist:	SAC	Date:	5 Jan 99

SAMPLE.WB1 REV 2.0

22010NML

4805

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	33862	GROSS WEIGHT	(W2)	7.4391
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3460
27743	BACKGROUND in cpm	(BKG)	4.5	NET WEIGHT	(W3)	0.0931
Test Code	SAMPLE VOLUME in mL	(SS)	0.500	DELTA TIME (HOURS)	(DT)	10.43
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.9992			
SOLID	SAMPLE COUNT RATE	(Rs)	3381.70	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.28	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
98006379	TIME OF SEPARATION	(ST)	10:54	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	12/31/98	DETECTION LIMIT	(Ld)	2.65
1	TIME OF COUNT	(TOC)	21:20	Sr-89/90 CONC in µCi/g 3.5002E+00		
Sample Prep	DATE OF COUNT	(DOC)	12/31/98			
FUSION01						
Sample #						
S98T002988						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
MCB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
CRJ	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
01/04/99						
Analysis Date						DETECTION LEVEL 2.74E-03 µCi/g
12/31/98	Sr-89/90 CONCENTRATION	3.50E+00	µCi/g			
Analysis Time	RELATIVE COUNTING ERROR	1.1%				
01:35 PM						
Sample Point						
BY-105	PERCENT CARRIER RECOVERY	93.1%				

Analyst:	CRJ	Date:	04-Jan-99
Signature of Chemist:	SAC	Date:	5 Jan 99

SAMPLE.WB1 REV 2.0

22010NML

4806

LABCORE Completed Worklist Report for Worklist# 27864

Analyst: gll Instrument: AB10 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY105 FOR @SR90-01 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01 SR90-01 SOLID		1.04E-03	1.03E-3	99.038 % Recovery	
1 STD	0	@SR90-01 SR90-01C SOLID		100	9.27E+01	92.700 % Recovery	
1 STD	0	@SR90-01 SR90-01E SOLID		1.00	2.05E+00	2.050 % Ct. Error	
2 BLNK-PREP	0	@SR90-01 SR90-01 SOLID		1	3.78E-2	0.038 uCi/g	
2 BLNK-PREP	0	@SR90-01 SR90-01C SOLID		100	9.21E+01	92.100 % Recovery	
2 BLNK-PREP	0	@SR90-01 SR90-01E SOLID		1.00	8.41E+00	8.410 uCi/g	
3 SAMPLE	S98T003393	0 F @SR90-01 SR90-01 SOLID		N/A	1.01E+00	2.02E-003 uCi/g	
3 SAMPLE	S98T003393	0 F @SR90-01 SR90-01C SOLID		N/A	9.14E+01	% Recovery	
3 SAMPLE	S98T003393	0 F @SR90-01 SR90-01E SOLID		N/A	1.46E+00	% Ct. Error	
4 DUP	S98T003393	0 F @SR90-01 SR90-01 SOLID		1.01E+0	8.55E-1	16.622 RPD	
4 DUP	S98T003393	0 F @SR90-01 SR90-01C SOLID		100	9.09E+01	90.900 % Recovery	
4 DUP	S98T003393	0 F @SR90-01 SR90-01E SOLID		1.00	1.59E+00	1.590 % Cnt Err	
5 SAMPLE	S98T003398	0 F @SR90-01 SR90-01 SOLID		N/A	6.84E+02	0.205 uCi/g	
5 SAMPLE	S98T003398	0 F @SR90-01 SR90-01C SOLID		N/A	9.18E+01	% Recovery	
5 SAMPLE	S98T003398	0 F @SR90-01 SR90-01E SOLID		N/A	5.64E-01	% Ct. Error	
6 DUP	S98T003398	0 F @SR90-01 SR90-01 SOLID		6.84E+2	6.89E+2	0.728 RPD	
6 DUP	S98T003398	0 F @SR90-01 SR90-01C SOLID		100	9.05E+01	90.500 % Recovery	
6 DUP	S98T003398	0 F @SR90-01 SR90-01E SOLID		1.00	5.63E-01	0.563 % Cnt Err	
7 SAMPLE	S98T003388	0 F @SR90-01 SR90-01 SOLID		N/A	1.42E+00	1.96E-003 uCi/g	
7 SAMPLE	S98T003388	0 F @SR90-01 SR90-01C SOLID		N/A	9.17E+01	% Recovery	
7 SAMPLE	S98T003388	0 F @SR90-01 SR90-01E SOLID		N/A	1.21E+00	% Ct. Error	
8 DUP	S98T003388	0 F @SR90-01 SR90-01 SOLID		1.42E+0	1.32E+0	7.299 RPD	
8 DUP	S98T003388	0 F @SR90-01 SR90-01C SOLID		100	9.20E+01	92.000 % Recovery	
8 DUP	S98T003388	0 F @SR90-01 SR90-01E SOLID		1.00	1.26E+00	1.260 % Cnt Err	

Final page for worklist# 27864

Analyst Signature _____ Date _____ Analyst Signature _____ Date _____

Reviewer Signature *John Relya* Date *15 Jan 99*

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 27864

Analyst: gll Instrument: AB00 10 Book# 2857

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105 FOR @SR90-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	SOLID		
2 BLNK-PREP			@SR90-01	SOLID		
3 SAMPLE	S98T003393	0 F	@SR90-01	SOLID	98000436	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
4 DUP	S98T003393	0 F	@SR90-01	SOLID		
5 SAMPLE	S98T003398	0 F	@SR90-01	SOLID	98000436	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
6 DUP	S98T003398	0 F	@SR90-01	SOLID		
7 SAMPLE	S98T003388	0 F	@SR90-01	SOLID	98000490	BY-105 (2)
Analytes Requested: SR90-01 , SR90-01C, SR90-01E						
8 DUP	S98T003388	0 F	@SR90-01	SOLID		

Final page for worklist # 27864

gll 1-14-99
Analyst Signature Date

En Baubi 1-15-99
Analyst Signature Date
MEB 1/15/99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD	
Type	DETECTOR NUMBER	10	CARRIER ADDED in mL (CVA)		1.000
STD	TOTAL COUNTS (TC)	9384	GROSS WEIGHT (W2)		7.4227
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)		7.3300
27864	BACKGROUND in cpm (BKG)	8.6	NET WEIGHT (W3)		0.0927
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)		6.75
@SR90-01	DILUTION FACTOR (DF)	1			
Matrix	DIGEST DILUTION FACTOR (DDF)	1			
LIQUID	SAMPLE COUNT RATE (Rs)	929.80	SR-90 EFFICIENCY FACTO (C1)		0.4051
Batch Number	CRITICAL LEVEL (Lc)	1.77	Y-90 EFFICIENCY FACTOR (C2)		0.4503
99000116	TIME OF SEPARATION (ST)	13:00	Rmax		N/A
Return	DATE OF SEPARATION (SD)	01/14/99	DETECTION LIMIT (Ld)		3.62
0	TIME OF COUNT (TOC)	19:45	Sr-89/90 CONC. in $\mu\text{Ci/L}$		1.0345E+00
Sample Prep	DATE OF COUNT (DOC)	01/14/99			
N/A	STANDARD BOOK #	28B57			
Sample	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0389E-03			
WL27864-STD					
Instrument Code					
WB26870	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS \geq 0$ OR REPLACE RS WITH Lc IF $RS < 0$				
EMB	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e \text{ to the power of } ((- \text{natural log } 2) / 64.2 * DT))) * SS * REC * 2220000)$				
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((W2 - W1) / (CVA * 0.1000))$				
SAC	Relative Counting Error = The Square Root of $((TC + BKG * CT) / (TC - BKG * CT)) * 1.96$				
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
GLL	NOTE: Expected weight = $CVA * 0.1$				
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
01/15/99	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
Analysis Date	Sr-89/90 CONCENTRATION				DETECTION LEVEL 4.03E-06 $\mu\text{Ci/L}$
01/14/99	1.03E-03 $\mu\text{Ci/mL}$				
Analysis Time	RELATIVE COUNTING ERROR				
03:00 PM	2.1%				
Sample Point	PERCENT CARRIER RECOVERY				
BY-105 (2)	92.7%				

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99

STANDARD.WB1 REV 2.0

22010NML

4809

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

BLNK-PREP

Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
BLNK-PREP	TOTAL COUNTS	(TC)	767	GROSS WEIGHT	(W2)	7.4066
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3145
27864	BACKGROUND in cpm	(BKG)	8.6	NET WEIGHT	(W3)	0.0921
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	6.95
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.014			
SOLID	SAMPLE COUNT RATE	(Rs)	68.10	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.77	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
99000116	TIME OF SEPARATION	(ST)	13:00	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	01/14/99	DETECTION LIMIT	(Ld)	3.62
0	TIME OF COUNT	(TOC)	19:57	Sr-89/90 CONC in $\mu\text{Ci/g}$		3.7787E-02
Sample/Rep	DATE OF COUNT	(DOC)	01/14/99			
N/A						
Sample						

WL27864 BLNK-PREP

Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)		
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0		
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))))*SS*(Dg/L)*REC*2220000		
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))		
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96		
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100		
Analyst	NOTE: Expected weight = CVA * 0.1		
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.		
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100		
01/15/99			
Analysis Date			
01/14/99	Sr-89/90 CONCENTRATION	3.78E-02	$\mu\text{Ci/g}$
Analysis Time			
03:00 PM	RELATIVE COUNTING ERROR	8.4%	
Sample Point			
BY-105 (2)	PERCENT CARRIER RECOVERY	92.1%	

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99
BLANK.WB1 REV 2.0	22010NML		

4810

WORKBOOK PAGE: SAM3

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)						SAMPLE
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	18259	GROSS WEIGHT	(W2)	7.3923
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3009
27864	BACKGROUND in cpm	(BKG)	8.6	NET WEIGHT	(W3)	0.0914
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	7.25
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L	2.014			
SOLID	SAMPLE COUNT RATE	(Rs)	1817.30	Sr-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.77	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
99000116	TIME OF SEPARATION	(ST)	13:00	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	01/14/99	DETECTION LIMIT	(Ld)	3.62
0	TIME OF COUNT	(TOC)	20:15	Sr-89/90 CONC in $\mu\text{Ci/g}$		1.0130E+00
Sample Prep	DATE OF COUNT	(DOC)	01/14/99			
FUSION01						
Sample #						
S98T003393						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} \geq 0$ or Replace RS with Lc if $\text{RS} < 0$					
Prepared By	$\text{RS} * 1000 * \text{DF} / ((\text{C1} + \text{C2} * (1 - e^{-(\ln(2) * \text{DT})})) * \text{SS} * (\text{Dg/L}) * \text{REC} * 2220000)$					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $(\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$					
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$					
01/15/99						
Analysis Date						
01/14/99	Sr-89/90 CONCENTRATION		1.01E+00	$\mu\text{Ci/g}$	DETECTION LEVEL	
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR		1.5%	2.02E-03 $\mu\text{Ci/g}$		
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY		91.4%			

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	<i>John Pelyea</i> 22010NML	SAC	Date: <i>15 Jan 99</i>

SAMPLE.WB1 REV 2.0

4811

WORKBOOK PAGE: DUP4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

				DUP	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC) 15538	GROSS WEIGHT	(W2) 7.4038	
Work Log	COUNT TIME in MINUTES	(CT) 10	TARE WEIGHT	(W1) 7.3129	
27864	BACKGROUND in cpm	(BKG) 8.6	NET WEIGHT	(W3) 0.0909	
Test Code	SAMPLE VOLUME in mL	(SS) 1.000	DELTA TIME (HOURS)	(DT) 7.67	
@SR90-01	DILUTION FACTOR	DF 1			
Matrix	DIGEST FACTOR (g/L)	(D g/L) 2.0312			
SOLID	SAMPLE COUNT RATE	(Rs) 1545.20	SR-90 EFFICIENCY FACTOR	(C1) 0.4051	
Batch Number	CRITICAL LEVEL	(Lc) 1.77	Y-90 EFFICIENCY FACTOR	(C2) 0.4503	
99000116	TIME OF SEPARATION	(ST) 13:00	Rmax	N/A	
Refun	DATE OF SEPARATION	(SD) 01/14/99	DETECTION LIMIT	(Ld) 3.62	
0	TIME OF COUNT	(TOC) 20:40	Sr-89/90 CONC in µCi/g	8.5507E-01	
Sample Prep	DATE OF COUNT	(DOC) 01/14/99			
FUSION01					
Sample #					
S98T003393					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in µCi/g Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*(Dg/L)*REC*2220000)				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
01/15/99					
Analysis Date					DETECTION LEVEL 2.01E-03 µCi/g
01/14/99	Sr-89/90 CONCENTRATION	8.55E-01	µCi/g		
Analysis Time					
03:00 PM	RELATIVE COUNTING ERROR	1.6%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	90.9%			

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99
SAMPLE.WB1 REV 2.0			

4812

WORKBOOK PAGE: SAM5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
SAMPLE	TOTAL COUNTS	(TC) 121036	GROSS WEIGHT	(W2) 7.4223	
Work List	COUNT TIME in MINUTES	(CT) 10	TARE WEIGHT	(W1) 7.3305	
27864	BACKGROUND in cpm	(BKG) 8.6	NET WEIGHT	(W3) 0.0918	
Test Code	SAMPLE VOLUME in mL	(SS) 1.000	DELTA TIME (HOURS)	(DT) 7.75	
@SR90-01	DILUTION FACTOR	DF 101			
Matrix	DIGEST FACTOR (g/L)	(D g/L) 1.986			
SOLID	SAMPLE COUNT RATE	(Rs) 12095.00	SR-90 EFFICIENCY FACTOR	(C1) 0.4051	
Batch Number	CRITICAL LEVEL	(Lc) 1.77	Y-90 EFFICIENCY FACTOR	(C2) 0.4503	
99000116	TIME OF SEPARATION	(ST) 13:05	Rmax	N/A	
Rerun	DATE OF SEPARATION	(SD) 01/14/99	DETECTION LIMIT	(Ld) 3.62	
0	TIME OF COUNT	(TOC) 20:50	Sr-89/90 CONC in $\mu\text{Ci/g}$	6.8403E+02	
Sample Prep	DATE OF COUNT	(DOC) 01/14/99			
FUSION01					
Sample #					
S98T003398					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $\text{RS} \leq \text{Lc}$ and $\text{RS} > 0$ or Replace RS with Lc if $\text{RS} < 0$				
Prepared By	$\text{RS} * 1000 * \text{DF} / ((\text{C1} + \text{C2} * (1 - e \text{ to the power of } ((-\text{natural log } 2) / 64.2 * \text{DT}))) * \text{SS} * (\text{Dg/L}) * \text{REC} * 2220000)$				
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery $((\text{W2} - \text{W1}) / (\text{CVA} * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(\text{TC} + \text{BKG} * \text{CT}) / (\text{TC} - \text{BKG} * \text{CT})$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $\text{CVA} * 0.1$				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((\text{DOC} - \text{SD}) * 24) + (\text{TOC} - \text{ST}) / 100$				
01/15/99					
Analysis Date					DETECTION LEVEL 2.05E-01 $\mu\text{Ci/g}$
01/14/99	Sr-89/90 CONCENTRATION	6.84E+02	$\mu\text{Ci/g}$		
Analysis Time					
03:00 PM	RELATIVE COUNTING ERROR	0.6%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	91.8%			

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99

SAMPLE.WB1 REV 2.0

22010JML

4813

WORKBOOK PAGE: DUP6

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA) 1.000
DUP	TOTAL COUNTS	(TC)	121274	GROSS WEIGHT	(W2) 7.4212
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1) 7.3307
27864	BACKGROUND in cpm	(BKG)	8.6	NET WEIGHT	(W3) 0.0905
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT) 8.77
@SR90-01	DILUTION FACTOR	DF	101		
Matrix	DIGEST FACTOR (g/L)	(D g/L)	1.984		
SOLID	SAMPLE COUNT RATE	(Rs)	12118.80	SR-90 EFFICIENCY FACTOR	(C1) 0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.77	Y-90 EFFICIENCY FACTOR	(C2) 0.4503
99000116	TIME OF SEPARATION	(ST)	13:05	Rmax	N/A
Return	DATE OF SEPARATION	(SD)	01/14/99	DETECTION LIMIT	(Ld) 3.62
0	TIME OF COUNT	(TOC)	21:51	Sr-89/90 CONC in $\mu\text{Ci/g}$	6.8886E+02
Sample Prep	DATE OF COUNT	(DOC)	01/14/99		
FUSION01					
Sample #					
S98T003398					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$				
Prepared By	$RS * 1000 * DF / ((C1 + C2 * (1 - e^{-(\ln(2) / (64.2 * DT))}) * SS * (Dg/L) * REC * 2220000)$				
EMB	NOTE: $64.2 = \text{Half Life for Y-90 and Rec.} = \text{Fractional Carrier Recovery} ((W2 - W1) / (CVA * 0.1000))$				
Chemist	Relative Counting Error = (The Square Root of $(TC + BKG * CT) / (TC - BKG * CT)$) * 1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = $CVA * 0.1$				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$				
01/15/99					
Analysis Date					
01/14/99	Sr-89/90 CONCENTRATION	6.89E+02		$\mu\text{Ci/g}$	DETECTION LEVEL 2.06E-01 $\mu\text{Ci/g}$
Analysis Time					
03:00 PM	RELATIVE COUNTING ERROR	0.6%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	90.5%			

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99
SAMPLE.WB1 REV 2.0			

4814

WORKBOOK PAGE: SAM7

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	26368	GROSS WEIGHT	(W2)	7.4194
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3277
27864	BACKGROUND in cpm	(BKG)	8.6	NET WEIGHT	(W3)	0.0917
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.70
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.022			
SOLID	SAMPLE COUNT RATE	(Rs)	2628.20	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.77	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
99000116	TIME OF SEPARATION	(ST)	13:05	Rmax		N/A
Rerun	DATE OF SEPARATION	(SD)	01/14/99	DETECTION LIMIT	(Ld)	3.62
0	TIME OF COUNT	(TOC)	22:47	Sr-89/90 CONC in $\mu\text{Ci/g}$		1.4193E+00
Sample Prep	DATE OF COUNT	(DOC)	01/14/99			
FUSION01						
Sample #						
S98T003388						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	RS*1000*DF/((C1+C2*(1-e to the power of ((-natural log 2)/(64.2*DT))))*SS*(Dg/L)*REC*2220000)					
EMB	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemis	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
01/15/99						
Analysis Date						DETECTION LEVEL 1.96E-03 $\mu\text{Ci/g}$
01/14/99	Sr-89/90 CONCENTRATION	1.42E+00		$\mu\text{Ci/g}$		
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	1.2%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	91.7%				

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99
SAMPLE.WB1 REV 2.0 22010NML			

4815

WORKBOOK PAGE: DUP8

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		10	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	24614	GROSS WEIGHT	(W2)	7.4060
Work List	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3140
27864	BACKGROUND in cpm	(BKG)	8.6	NET WEIGHT	(W3)	0.0920
Test Code	SAMPLE VOLUME in mL	(SS)	1.000	DELTA TIME (HOURS)	(DT)	9.87
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST FACTOR (g/L)	(D g/L)	2.024			
SOLID	SAMPLE COUNT RATE	(Rs)	2452.80	SR-90 EFFICIENCY FACTOR	(C1)	0.4051
Batch Number	CRITICAL LEVEL	(Lc)	1.77	Y-90 EFFICIENCY FACTOR	(C2)	0.4503
99000116	TIME OF SEPARATION	(ST)	13:05	Rmax		N/A
Recun	DATE OF SEPARATION	(SD)	01/14/99	DETECTION LIMIT	(Ld)	3.62
0	TIME OF COUNT	(TOC)	22:57	Sr-89/90 CONC in $\mu\text{Ci/g}$		1.3168E+00
Sample Prep	DATE OF COUNT	(DOC)	01/14/99			
FUSION01						
Sample #						
S98T003388						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB26870	Sr-89/90 CONC in $\mu\text{Ci/g}$ Replace RS with RMAX if $RS \leq Lc$ and $RS > 0$ or Replace RS with Lc if $RS < 0$					
Prepared By	$RS * 1000 * DF / ((C1 + C2 * (1 - e^{-(\ln(2) * DT) / 64.2 * DT)})) * SS * (Dg/L) * REC * 2220000$					
EMB	NOTE: $64.2 = \text{Half Life for Y-90 and Rec.} = \text{Fractional Carrier Recovery } ((W2 - W1) / (CVA * 0.1000))$					
Chemist	Relative Counting Error = $(\text{The Square Root of } (TC + BKG * CT) / (TC - BKG * CT)) * 1.96$					
SAC	Percent Carrier Recovery = $(\text{Net Weight} / \text{Expected weight}) * 100$					
Analyst	NOTE: Expected weight = $CVA * 0.1$					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = $((DOC - SD) * 24) + (TOC - ST) / 100$					
01/15/99						
Analysis Date						DETECTION LEVEL
01/14/99	Sr-89/90 CONCENTRATION	1.32E+00			$\mu\text{Ci/g}$	
Analysis Time						
03:00 PM	RELATIVE COUNTING ERROR	1.3%				
Sample Point						1.95E-03
BY-105 (2)	PERCENT CARRIER RECOVERY	92.0%				$\mu\text{Ci/g}$

Analyst:	GLL	Date:	15-Jan-99
Signature of Chemist:	SAC	Date:	15 Jan 99
SAMPLE.WB1 REV 2.0			

4816

LABCORE Completed Worklist Report for Worklist# 28060

Analyst: gll Instrument: AB12 Book#: _____

Method: LA-220-101 Rev/Mod _____

Worklist Comment: BY105 FIELD BLANK FOR @SR90-01 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@SR90-01	SR90-01	LIQUID	1.04E-03	1.03E-3	99.038 %	Recovery
1 STD	0	@SR90-01	SR90-01C	LIQUID	100	9.36E+01	93.600 %	Recovery
1 STD	0	@SR90-01	SR90-01E	LIQUID	1.00	1.96E+00	1.960 %	Ct. Err
2 BLNK	0	@SR90-01	SR90-01	LIQUID	1	<2.08E-7		uCi/mL
2 BLNK	0	@SR90-01	SR90-01C	LIQUID	100	9.32E+01	93.200 %	Recovery
2 BLNK	0	@SR90-01	SR90-01E	LIQUID	1.00	4.85E+02	485.000	uCi/mL
3 BLNK/BKG	0	@SR90-01	SR90-01	LIQUID	1	8.97E-01	0.897	BLNK/BKG
4 SAMPLE	S98T002391	0	@SR90-01	SR90-01	LIQUID	N/A	<	3.23E-07
4 SAMPLE	S98T002391	0	@SR90-01	SR90-01C	LIQUID	N/A		9.31E+01
4 SAMPLE	S98T002391	0	@SR90-01	SR90-01E	LIQUID	N/A		3.11E+02
5 DUP	S98T002391	0	@SR90-01	SR90-01	LIQUID	<3.23E-7		<4.13E-7
5 DUP	S98T002391	0	@SR90-01	SR90-01C	LIQUID	100	9.31E+01	93.100 %
5 DUP	S98T002391	0	@SR90-01	SR90-01E	LIQUID	1.00	1.78E+02	178.000 %

Final page for worklist# 28060

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

3 Feb 99
Reviewer Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 28060

Analyst: gll Instrument: AB00 12 Book# 28B57

Method: LA-220-101 Rev/Mod E-3

Worklist Comment: BY105 FIELD BLANK FOR @SR90-01 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@SR90-01	LIQUID		
2 BLNK			@SR90-01	LIQUID		
3 BLNK/BKG			@SR90-01	LIQUID		
4 SAMPLE	S98T002391 0		@SR90-01	LIQUID	98000436	BY-105 (2)
	Analytes Requested: SR90-01 , SR90-01C, SR90-01E					
5 DUP	S98T002391 0		@SR90-01	LIQUID		

Final page for worklist # 28060

gll 2-2-99
Analyst Signature Date

Supra 2 Chan 2/3/99
Analyst Signature Date
2/3/99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5) LIQUIDS

				STANDARD
Type	DETECTOR NUMBER	12	CARRIER ADDED in mL (CVA)	1.000
STD	TOTAL COUNTS (TC)	10126	GROSS WEIGHT (W2)	7.4256
Work List	COUNT TIME in MINUTES (CT)	10	TARE WEIGHT (W1)	7.3320
28060	BACKGROUND in cpm (BKG)	2.9	NET WEIGHT (W3)	0.0936
Test Code	SAMPLE VOLUME in mL (SS)	1.000	DELTA TIME (HOURS) (DT)	10.87
@SR90-01	DILUTION FACTOR (DF)	1		
Matrix	DIGEST DILUTION FACTOR (DDF)	1		
LIQUID	SAMPLE COUNT RATE (Rs)	1009.70	SR-90 EFFICIENCY FACTO (C1)	0.4180
Batch Number	CRITICAL LEVEL (Lc)	1.03	Y-90 EFFICIENCY FACTOR (C2)	0.4660
99000332	TIME OF SEPARATION (ST)	09:55	Rmax	N/A
Re-run	DATE OF SEPARATION (SD)	02/02/99	DETECTION LIMIT (Ld)	2.14
0	TIME OF COUNT (TOC)	20:47	Sr-89/90 CONC. in $\mu\text{Ci/L}$	1.0348E+00
Sample Prep	DATE OF COUNT (DOC)	02/02/99		
N/A	STANDARD BOOK #	28B57		
Sample #	STANDARD VALUE in $\mu\text{Ci/mL}$	1.0376E-03		
WL28060				
Instrument Code				
WB27811	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)			
Prepared By	Sr-89/90 CONC in $\mu\text{Ci/mL}$ REPLACE RS WITH RMAX IF $RS \leq Lc$ AND $RS > 0$ OR REPLACE RS WITH Lc IF $RS < 0$			
SZC	$RS * DF * DDF * 1000 / ((C1 + C2 * (1 - e^{-((1 - \text{natural log } 2) / 64.2 * DT))}) * SS * REC * 2220000)$			
Chemist	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))			
SAC	Relative Counting Error = The Square Root of ((TC + BKG * CT) / (TC - BKG * CT) * 1.96)			
Analyst	Percent Carrier Recovery = (Net Weight / Expected weight) * 100			
GLL	NOTE: Expected weight = CVA * 0.1			
Date Complete	Detection Levels and Less Than Values are determined from Procedure LA-508-002.			
02/03/99	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100			
Analysis Date				DETECTION LEVEL 2.20E-06 $\mu\text{Ci/L}$
02/02/99	Sr-89/90 CONCENTRATION	1.03E-03	$\mu\text{Ci/mL}$	
Analysis Time				
01:10 PM	RELATIVE COUNTING ERROR	2.0%		
Sample Point				
BY-105 (2)	PERCENT CARRIER RECOVERY	93.6%		

Analyst:	GLL	Date:	03-Feb-99
Signature of Chemist:		SAC	Date: 3 Feb 99

STANDARD.WB1 REV 2.0

22010NML

4819

WORKBOOK PAGE: BLANK2

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

LA-220-101 / E-3 Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)				BLNK	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL (CVA)	1.000
BLNK	TOTAL COUNTS (TC)		26	GROSS WEIGHT (W2)	7.4218
Worklist	COUNT TIME in MINUTES (CT)		10	TARE WEIGHT (W1)	7.3286
28060	BACKGROUND in cpm (BKG)		2.9	NET WEIGHT (W3)	0.0932
Test Code	SAMPLE VOLUME in mL (SS)		5.000	DELTA TIME (HOURS) (DT)	12.30
@SR90-01	DILUTION FACTOR (DF)		1		
Matrix	DIGEST DILUTION FACTOR (DDF)		1		
LIQUID	SAMPLE COUNT RATE (Rs)	-0.30	SR-90 EFFICIENCY FACTOR (C1)		0.4180
Batch Number	CRITICAL LEVEL (Lc)	1.03	Y-90 EFFICIENCY FACTOR (C2)		0.4660
99000332	TIME OF SEPARATION (ST)	09:55	Rmax		N/A
ReRun	DATE OF SEPARATION (SD)	02/02/99	DETECTION LIMIT (Ld)		2.14
0	TIME OF COUNT (TOC)	22:13	Sr-89/90 CONC in µCi/L	<	2.0838E-04
Sample Prep	DATE OF COUNT (DOC)	02/02/99			
N/A					
Sample #					
WL28060					
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)				
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0				
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))**SS*REC*2220000)				
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))				
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96				
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100				
Analyst	NOTE: Expected weight = CVA * 0.1				
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.				
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100				
02/03/99					
Analysis Date					
02/02/99	Sr-89/90 CONCENTRATION	< 2.08E-07	µCi/mL	DETECTION LEVEL 4.35E-07 µCi/mL	
Analysis Time	LESS THAN Value was Determined from Lc.				
01:10 PM	RELATIVE COUNTING ERROR	484.5%			
Sample Point					
BY-105 (2)	PERCENT CARRIER RECOVERY	93.2%			

Analyst:	GLL	Date:	03-Feb-99
Signature of Chemist:	SAC	Date:	3 Feb 99

BLANK.WB1 REV 2.0

22010NML

WORKBOOK PAGE: SAM4

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					SAMPLE	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
SAMPLE	TOTAL COUNTS	(TC)	34	GROSS WEIGHT	(W2)	7.3857
Worklist	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.2926
28060	BACKGROUND in cpm	(BKG)	2.9	NET WEIGHT	(W3)	0.0931
Test Code	SAMPLE VOLUME in mL	(SS)	5.000	DELTA TIME (HOURS)	(DT)	12.50
@SR90-01	DILUTION FACTOR	DF	1			
Matrix	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	0.50	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.03	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
99000332	TIME OF SEPARATION	(ST)	09:55	Rmax		1.59
ReRun	DATE OF SEPARATION	(SD)	02/02/99	DETECTION LIMIT	(Ld)	2.14
0	TIME OF COUNT	(TOC)	22:25	Sr-89/90 CONC in µCi/L	<	3.2274E-04
Sample Prep	DATE OF COUNT	(DOC)	02/02/99			
N/A						
Sample #						
S98T002391						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT)))*SS*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemist	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT))*1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
02/03/99						
Analysis Date						DETECTION LEVEL 4.35E-07 µCi/mL
02/02/99	Sr-89/90 CONCENTRATION	< 3.23E-07	µCi/mL			
Analysis Time	LESS THAN Value was Determined from Rmax.					
01:10 PM	RELATIVE COUNTING ERROR	311.1%				
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY	93.1%				

Analyst:	GLL	Date:	03-Feb-99
Signature of Chemist:	SAC	Date:	3 Feb 99

SAMPLE.WB1 REV 2.0

22010NML

4821

WORKBOOK PAGE: DUP5

Sr-89/90 : LA-220-101 (E-3), 102 (INACTIVE), 104 (E-5)

					DUP	
Type	DETECTOR NUMBER		12	CARRIER ADDED in mL	(CVA)	1.000
DUP	TOTAL COUNTS	(TC)	38	GROSS WEIGHT	(W2)	7.4192
WorkList	COUNT TIME in MINUTES	(CT)	10	TARE WEIGHT	(W1)	7.3261
28060	BACKGROUND in cpm	(BKG)	2.9	NET WEIGHT	(W3)	0.0931
Test Code	SAMPLE VOLUME in mL	(SS)	5.000	DELTA TIME (HOURS)	(DT)	12.68
@SR90-01	DILUTION FACTOR	DF	1			
Matrx	DIGEST DILUTION FACTOR	(DDF)	1			
LIQUID	SAMPLE COUNT RATE	(Rs)	0.90	SR-90 EFFICIENCY FACTOR	(C1)	0.4180
Batch Number	CRITICAL LEVEL	(Lc)	1.03	Y-90 EFFICIENCY FACTOR	(C2)	0.4660
99000332	TIME OF SEPARATION	(ST)	09:55	Rmax		2.04
Rerun	DATE OF SEPARATION	(SD)	02/02/99	DETECTION LIMIT	(Ld)	2.14
0	TIME OF COUNT	(TOC)	22:36	Sr-89/90 CONC in µCi/L	<	4.1313E-04
Sample Prep	DATE OF COUNT	(DOC)	02/02/99			
N/A						
Sample #						
S98T002391						
Instrument Code	Sample Count Rate (Rs) = (Total Counts (TC) / Count Time (CT)) - Background in cpm (BKG)					
WB27811	Sr-89/90 CONC in µCi/L Replace RS with RMAX if RS<=Lc and RS>=0 or Replace RS with Lc if RS<0					
Prepared By	RS*DF*DDF/((C1+C2*(1-e to the power of ((-natural log 2)/64.2*DT))) * SS*REC*2220000)					
SZC	NOTE: 64.2 = Half Life for Y-90 and Rec. = Fractional Carrier Recovery ((W2-W1) / (CVA * 0.1000))					
Chemists	Relative Counting Error = (The Square Root of (TC + BKG * CT) / (TC - BKG * CT)) * 1.96					
SAC	Percent Carrier Recovery = (Net Weight / Expected weight) * 100					
Analyst	NOTE: Expected weight = CVA * 0.1					
GLL	Detection Levels and Less Than Values are determined from Procedure LA-508-002.					
Date Complete	Delta Time (hours) = ((DOC - SD) * 24) + (TOC - ST) / 100					
02/03/99						
Analysis Date						
02/02/99	Sr-89/90 CONCENTRATION < 4.13E-07 µCi/mL					DETECTION LEVEL 4.34E-07 µCi/mL
Analysis Time	LESS THAN Value was Determined from Rmax.					
01:10 PM	RELATIVE COUNTING ERROR 178.3%					
Sample Point						
BY-105 (2)	PERCENT CARRIER RECOVERY 93.1%					

Analyst:	GLL	Date:	03-Feb-99
Signature of Chemist:	SAC	Date:	3 Feb 99
SAMPLE.WB1 REV 2.0		22010NML	

4822

**This document was too large to scan
as a single document; therefore, it has
been divided into smaller sections.**

Section 17 of 17

Document Information			
Document #	HNF-1663	Revision	0
Title	TANK 241BY105 CORES 246 & 246R & 250 & 251 & 252 ANALYTICAL RESULTS FOR THE FINAL REPORT (PAGES 4823 THRU 5113)		
Date	4/6/1999		
Originator	STEEN FH	Originator Co.	WMH
References	EDT-626135		
Keywords	TANK CHARACTERIZATION		
Other Information	PLUTONIUM 239 ANALYSIS, PU239 WORKLIST, OPPORTUNISTIC ANALYTES, TRADEMARKS		

HNF-1663 REV. 0
TABLE OF CONTENTS

Sec 1
of 17

Narrative	1
BY-105 Sample Breakdown (Attachment 1)	31
Correspondence (Attachment 2)	75
Sample Data Summary	79
Chain of Custody Forms	161
Photographs	201
Sample Handling	239
Core Composite Worksheets	241
Extrusion Worklist # 25629 (2475)	256
Extrusion Worklist # 25630 (2476)	257
Extrusion Worklist # 25631 (2472)	258
Extrusion Worklist # 25632 (2477)	259
Extrusion Worklist # 25633 (2473)	260
Extrusion Worklist # 25634 (2474)	261
Extrusion Worklist # 25798 (2625)	262
Extrusion Worklist # 25803 (2626)	263
Extrusion Worklist # 25805 (2627)	264
Extrusion Worklist # 25807 (2628)	265
Extrusion Worklist # 25808 (2629)	266
Extrusion Worklist # 25809 (2630)	267
Extrusion Worklist # 25811 (2631)	268
Extrusion Worklist # 25814 (2632)	269
Extrusion Worklist # 25815 (2633)	270
Extrusion Worklist # 25830 (2634)	271
Extrusion Worklist # 25831 (2635)	272
Extrusion Worklist # 25832 (2636)	273
Extrusion Worklist # 26100 (2725)	274
Extrusion Worklist # 26102 (2726)	275
Extrusion Worklist # 26116 (2727)	276
Extrusion Worklist # 26117 (2728)	277
Extrusion Worklist # 26118 (2729)	278
Extrusion Worklist # 26119 (2730)	279
Extrusion Worklist # 26142 (2719)	280
Extrusion Worklist # 26145 (2720)	281
Extrusion Worklist # 26146 (2721)	282
Extrusion Worklist # 26147 (2722)	283
Extrusion Worklist # 26148 (2723)	284
Extrusion Worklist # 26217 (2850)	285

TABLE OF CONTENTS (Continued)

Extrusion Worklist # 26149 (2724)	286
Extrusion Worklist # 26150 (2731)	287
Extrusion Worklist # 26151 (2732)	288
Extrusion Worklist # 26152 (2733)	289
Extrusion Worklist # 26153 (2734)	290
Extrusion Worklist # 26155 (2735)	291
Sample Preparations	292
Fusion Digestion Worklist # 26816 (2598, 2599, 2600)	294
Fusion Digestion Worklist # 26817 (2601, 2602, 2603)	295
Fusion Digestion Worklist # 26818 (2712, 2739, 2844)	296
Fusion Digestion Worklist # 26819 (2845, 2918, 2919)	297
Fusion Digestion Worklist # 26820 (2920, 2921, 2922)	298
Fusion Digestion Worklist # 26821 (2986, 2987)	299
Fusion Digestion Worklist # 26822 (2988, 2989)	300
Fusion Digestion Worklist # 27651 (3388, 3393, 3398)	301
H2O Digestion Worklist # 26824 (2610, 2611, 2612)	302
H2O Digestion Worklist # 26826 (2613, 2614, 2615)	303
H2O Digestion Worklist # 26828 (2714, 2741, 2848)	304
H2O Digestion Worklist # 26830 (2849, 2931, 2932)	305
H2O Digestion Worklist # 26832 (2933, 2934, 2935)	306
H2O Digestion Worklist # 26834 (2994, 2995, 2996, 2997)	307
H2O Digestion Worklist # 27653 (3390, 3395, 3400)	309
Acid Digestion Worklist # 26823 (2604, 2605, 2606)	310
Acid Digestion Worklist # 26825 (2607, 2608, 2609)	312
Acid Digestion Worklist # 26827 (2713, 2740, 2846)	314
Acid Digestion Worklist # 26829 (2847, 2926, 2927)	316
Acid Digestion Worklist # 26831 (2928, 2929, 2930)	318
Acid Digestion Worklist # 26833 (2990, 2991, 2992, 2993)	320
Acid Digestion Worklist # 27526 (3350, 3351)	322
Acid Digestion Worklist # 27652 (3389, 3394, 3399)	323
BY-105 Bulk Density Worksheets	325
Inorganic Analysis	348
Differential Scanning Calorimetry Analysis (DSC)	
DSC Worklist # 26781 (2806)	350
DSC Worklist # 26782 (2807, 2808)	354
DSC Worklist # 26783 (2809, 2810)	360
DSC Worklist # 26784 (2812)	366
DSC Worklist # 26785 (2813, 2938)	370
DSC Worklist # 26786 (2957, 2958, 2959)	376
DSC Worklist # 26787 (2955, 2956)	385

TABLE OF CONTENTS (Continued)

DSC Worklist # 26788 (2589, 2590)	391
DSC Worklist # 26789 (2591, 2592)	398
DSC Worklist # 26790 (2593, 2594)	404
DSC Worklist # 26791 (2711, 2737)	410
DSC Worklist # 26792 (2840, 2841)	416
DSC Worklist # 26793 (2910, 2911)	422
DSC Worklist # 26794 (2912, 2913)	428
DSC Worklist # 26795 (2914, 2980)	434
DSC Worklist # 26796 (2981, 2982, 2983)	440
DSC Worklist # 26852 (3002)	449
DSC Worklist # 26854 (3004, 3005)	456
DSC Worklist # 26916 (2957, 2958, 2959)	462
DSC Worklist # 26917 (2589, 2590, 2840, 2910, 2911, 2912)	463
DSC Worklist # 26918 (2913, 2914, 2980, 2981, 2982, 2983)	464
DSC Worklist # 27166 (2811)	465
DSC Worklist # 27173 (2806, 2807, 2808, 2813, 2938, 2955 2956, 3002, 3003, 2809, 2810)	469
DSC Worklist # 27174 (2591, 2592, 2593, 2594, 2711, 2737)	471
DSC Worklist # 27230 (2811, 2812, 3004, 3005)	472
DSC Worklist # 27231 (2841)	473
DSC Worklist # 27633 (3383)	474
DSC Worklist # 27634 (3191, 3396)	479
DSC Worklist # 27643 (3386)	486
DSC Worklist # 27765 (3386, 3391, 3396)	491
DSC Worklist # 27811 (3383)	492
DSC Worklist # 28062 (2391)	493
DSC Worklist # 28888 (2391)	498
Thermogravimetric Analysis (TGA)	
TGA Worklist # 26765 (2806)	499
TGA Worklist # 26766 (2807, 2808)	503
TGA Worklist # 26767 (2809, 2810)	509
TGA Worklist # 26768 (2811, 2812)	515
TGA Worklist # 26769 (2813, 2938)	521
TGA Worklist # 26770 (2955, 2956)	527
TGA Worklist # 26771 (2957, 2958, 2959)	533
TGA Worklist # 26772 (2589, 2590)	541
TGA Worklist # 26773 (2591, 2592)	547
TGA Worklist # 26774 (2593, 2594)	553
TGA Worklist # 26775 (2711, 2737)	559
TGA Worklist # 26776 (2840, 2841)	565
TGA Worklist # 26777 (2910, 2911)	571
TGA Worklist # 26778 (2912, 2913)	581
TGA Worklist # 26779 (2914, 2980)	587
TGA Worklist # 26780 (2981, 2982, 2983)	593
TGA Worklist # 26851 (3002, 3003)	602

TABLE OF CONTENTS (Continued)

TGA Worklist # 26853 (3004, 3005)	608
TGA Worklist # 26912 (2841)	614
TGA Worklist # 27637 (3383)	618
TGA Worklist # 27638 (3391, 3396)	623
TGA Worklist # 27644 (3386)	630
TGA Worklist # 28072 (2391)	635

Specific Gravity Analysis (SpG)

SpG Worklist # 28183 (2806, 2807, 2808, 2809)	640
SpG Worklist # 28184 (2810, 2811, 2812, 2813)	651
SpG Worklist # 28185 (2938, 2955, 2956, 2957)	662
SpG Worklist # 28186 (2958, 2959, 3002, 3003)	673
SpG Worklist # 28187 (3004, 3005, 3384)	684

Hydroxide Analysis (OH)

OH Worklist # 26731 (2822, 2823, 2824, 2825)	693
OH Worklist # 26732 (2826, 2827, 2828, 2829, 2940)	709
OH Worklist # 26733 (2965, 2966, 2967, 2968, 2969)	727
OH Worklist # 26843 (3010, 3011, 3012, 3013)	745
OH Worklist # 27642 (3385)	761
OH Worklist # 28070 (2391)	768
OH Worklist # 28138 (3390, 3395, 3400)	775

Ammonia Analysis (NH3)

NH3 Worklist # 26728 (2806, 2807)	787
NH3 Worklist # 26730 (2955)	812
NH3 Worklist # 26842 (3002, 3003)	826
NH3 Worklist # 26882 (2956, 2957)	842
NH3 Worklist # 26883 (2958, 2959)	859
NH3 Worklist # 26885 (2812, 2813)	877
NH3 Worklist # 26886 (2938)	893
NH3 Worklist # 26887 (3004)	907
NH3 Worklist # 26890 (3808, 3809)	921
NH3 Worklist # 27640 (3383)	937
NH3 Worklist # 27835 (3005)	950
NH3 Worklist # 27849 (2811, 2810)	962
NH3 Worklist # 28068 (2391)	980
NH3 Worklist # 28749 (2810, 2811)	992

Chromium VI Analysis (CRVI)

CRVI Worklist # 26739 (2822, 2823)	1010
CRVI Worklist # 26740 (2824, 2825, 2826)	1016
CRVI Worklist # 26741 (2827, 2828, 2829)	1023
CRVI Worklist # 26742 (2940, 2965, 2966)	1031
CRVI Worklist # 26743 (2967)	1040
CRVI Worklist # 26844 (3010, 3011)	1046
CRVI Worklist # 26845 (3012, 3013)	1052
CRVI Worklist # 26984 (2968, 2969)	1059

TABLE OF CONTENTS (Continued)

CRVI Worklist # 27639 (3385)	1066
CRVI Worklist # 27689 (3390, 3395, 3400)	1072
CRVI Worklist # 27737 (2934, 2935)	1084
CRVI Worklist # 28075 (2391)	1091

Ion Chromatographic Analysis (IC)	
IC Worklist # 25891 (2391, 2392)	1096
IC Worklist # 26756 (2814, 2815, 2816)	1106
IC Worklist # 26757 (2817, 2818, 2819)	1118
IC Worklist # 26758 (2820, 2821, 2960, 2961)	1131
IC Worklist # 26759 (2962, 2963)	1147
IC Worklist # 26760 (2391)	1159
IC Worklist # 26761 (2816, 2817, 2818)	1168
IC Worklist # 26762 (2819, 2820, 2821)	1187
IC Worklist # 26847 (3006, 3007)	1200
IC Worklist # 26848 (3008, 3009)	1210
IC Worklist # 26849 (3007)	1221
IC Worklist # 26850 (3008, 3009)	1229
IC Worklist # 26988 (2964)	1243
IC Worklist # 27021 (2960, 2961)	1251
IC Worklist # 27024 (2963)	1263
IC Worklist # 27276 (2994, 2995, 2996, 2997)	1273
IC Worklist # 27277 (2994, 2995, 2996, 2997)	1292
IC Worklist # 27291 (2994, 2996)	1315
IC Worklist # 27292 (2995, 2297)	1332
IC Worklist # 27395 (2610, 2611, 2612)	1349
IC Worklist # 27396 (2610, 2611, 2612)	1364
IC Worklist # 27635 (3384)	1378
IC Worklist # 27636 (3384)	1392
IC Worklist # 27686 (2714, 2741)	1401
IC Worklist # 27687 (3390, 3395, 3400)	1414
IC Worklist # 27692 (3390, 3395)	1432
IC Worklist # 27693 (2714, 2741, 2848)	1450
IC Worklist # 27732 (2612, 2614)	1466
IC Worklist # 27733 (2849, 2931, 2932)	1478
IC Worklist # 27734 (2933, 2934)	1492
IC Worklist # 27735 (2849, 2932)	1505
IC Worklist # 27736 (2933, 2934, 2935)	1522
IC Worklist # 27751 (3384)	1542
IC Worklist # 27761 (2613, 2615)	1552
IC Worklist # 27763 (2613)	1563
IC Worklist # 27850 (3400)	1572
IC Worklist # 27859 (2994, 2995, 2996)	1584
IC Worklist # 27885 (2814)	1600
IC Worklist # 27933 (3006)	1610
IC Worklist # 27935 (2934, 2935)	1617
IC Worklist # 27991 (2931)	1631

See 3 of
17
ContinuedSee 4
of 17See 5
of 17

TABLE OF CONTENTS (Continued)

IC Worklist # 28021 (2613, 2615)	1640
IC Worklist # 28066 (3390, 3400)	1651
IC Worklist # 28251 (2612)	1664
IC Worklist # 28270 (2849, 2931)	1675
IC Worklist # 28318 (2848, 2613)	1686
IC Worklist # 28341 (2714)	1700
IC Worklist # 28346 (2849, 2932)	1709
IC Worklist # 28347 (2714, 2741)	1721
IC Worklist # 28349 (2939, 2964)	1735
IC Worklist # 28394 (2933, 2934, 2935)	1746
IC Worklist # 28395 (2960, 2961, 2962)	1763
IC Worklist # 28406 (2848)	1780
IC Worklist # 28476 (2614, 2615)	1788
IC Worklist # 28541 (2932)	1800
IC Worklist # 28542 (2613)	1810
IC Worklist # 28648 (2815)	1818
IC Worklist # 28718 (2614, 2615)	1827

Inductively Coupled Plasma Analysis (ICP)

ICP Worklist # 25720 (2391, 2392)	1839
ICP Worklist # 26691 (2814, 2815, 2819)	1845
ICP Worklist # 26693 (2816, 2817, 2939)	1852
ICP Worklist # 26694 (2818, 2820, 2960)	1859
ICP Worklist # 26695 (2821, 2961, 2962)	1866
ICP Worklist # 26892 (2963, 2964)	1873
ICP Worklist # 26893 (3006, 3007, 3008, 3009)	1879
ICP Worklist # 26896 (2598, 2599, 2600)	1886
ICP Worklist # 26898 (2601, 2602, 2603)	1894
ICP Worklist # 26947 (2845, 2918, 2919)	1902
ICP Worklist # 27097 (2844, 2739, 2712)	1909
ICP Worklist # 27099 (2920, 2921, 2922)	1917
ICP Worklist # 27100 (2986, 2987)	1925
ICP Worklist # 27101 (2988, 2989)	1932
ICP Worklist # 27268 (2990, 2991, 2992, 2993)	1938
ICP Worklist # 27442 (2604, 2605, 2606)	1949
ICP Worklist # 27672 (2713, 2846, 2740)	1956
ICP Worklist # 27674 (3394, 3399, 3389)	1965
ICP Worklist # 27676 (3384, 3384)	1976
ICP Worklist # 27696 (2928, 2929, 2930)	1982
ICP Worklist # 27738 (2607, 2608, 2609)	1990
ICP Worklist # 27739 (2847, 2926, 2927)	1998
ICP Worklist # 27861 (3393, 3398, 3388)	2006

Inductive Coupled Plasma / Mass Spectrometer Analysis (ICP/MS)

ICP/MS Worklist # 26708 (2822, 2823, 2824, 2825, 2827, 2940)	2015
ICP/MS Worklist # 26709 (2826, 2828, 2829, 3385)	2074
ICP/MS Worklist # 26710 (2968, 2969, 3010, 3011)	2108
ICP/MS Worklist # 27269 (2990, 2991, 2992, 2993)	2150

TABLE OF CONTENTS (Continued)

ICP/MS Worklist # 27270	(2990, 2991, 2992, 2993)	2199
ICP/MS Worklist # 27475	(3012, 3013)	2233
ICP/MS Worklist # 27492	(2604, 2605, 2606)	2261
ICP/MS Worklist # 27493	(2604, 2605, 2606)	2286
ICP/MS Worklist # 27494	(2822, 2823, 2827)	2311
ICP/MS Worklist # 27495	(2822, 2823, 2827)	2329
ICP/MS Worklist # 27496	(2824, 2825, 2940)	2359
ICP/MS Worklist # 27497	(2824, 2825, 2940)	2380
ICP/MS Worklist # 27498	(2826, 2828, 2965, 3385, 2391)	2402
ICP/MS Worklist # 27499	(2826, 2828, 2965)	2429
ICP/MS Worklist # 27500	(2829, 2966, 2967)	2445
ICP/MS Worklist # 27501	(2829, 2966, 2967)	2463
ICP/MS Worklist # 27502	(2968, 2969, 3010)	2478
ICP/MS Worklist # 27503	(2968, 2969, 3010)	2499
ICP/MS Worklist # 27504	(3011, 3012, 3013)	2531
ICP/MS Worklist # 27505	(3011, 3012, 3013)	2552
ICP/MS Worklist # 27507	(2990, 2991, 2992, 2993)	2584
ICP/MS Worklist # 27699	(3394, 3389)	2611
ICP/MS Worklist # 27701	(3385)	2637
ICP/MS Worklist # 27705	(2713, 2846, 2740)	2654
ICP/MS Worklist # 27706	(2928, 2929)	2683
ICP/MS Worklist # 27707	(3394, 3399, 3389)	2708
ICP/MS Worklist # 27714	(2713, 2846, 2740)	2748
ICP/MS Worklist # 27715	(2928, 2929, 2930)	2788
ICP/MS Worklist # 27716	(2740)	2824
ICP/MS Worklist # 27717	(2928, 2929, 2930)	2848
ICP/MS Worklist # 27718	(3394, 3399, 3389, 2607, 2608, 2609)	2863
ICP/MS Worklist # 27741	(2847, 2926, 2927)	2901
ICP/MS Worklist # 27742	(2847, 2926, 2927)	2936
ICP/MS Worklist # 27744	(2847, 2926, 2927)	2962
ICP/MS Worklist # 27745	(2607, 2608, 2609)	2988
ICP/MS Worklist # 27746	(2713, 2846)	3016
ICP/MS Worklist # 27946	(2604, 2605, 2606)	3037
ICP/MS Worklist # 27947	(2607, 2608, 2609)	3070
ICP/MS Worklist # 28434	(3399, 2930)	3101
ICP/MS Worklist # 28448	(2965, 2966, 2967)	3129
ICP/MS Worklist # 28459	(2391)	3166
ICP/MS Worklist # 28488	(2391)	3182
ICP/MS Worklist # 28544	(2391, 2822, 2823, 2824, 2826, 2827, 2828, 2829, 2817)	3204
ICP/MS Worklist # 28545	(2940, 2965, 2966, 2969, 3010, 3011, 3012, 3012, 3013, 3385)	3239
ICP/MS Worklist # 28546	(2604, 2605, 2606)	3269
ICP/MS Worklist # 28547	(2713, 2740, 2846, 2847, 2926, 2927)	3288
ICP/MS Worklist # 28557	(2928, 2929, 2930, 2990, 2991, 2992, 2993)	3317
ICP/MS Worklist # 28558	(3389, 3394, 3399)	3348
ICP/MS Worklist # 28845	(2809, 2968)	3367
ICP/MS Worklist # 28912	(2607, 2608, 2909)	3382

Sec 7
of 17
Cont.

Sec 8
of 17

Sec 9
of 17

Sec 10
of 17

TABLE OF CONTENTS (Continued)

Total Inorganic Carbon/Total Organic Carbon Analysis (TICTOC)	
TICTOC Worklist # 26744 (2822, 2823)	3401
TICTOC Worklist # 26745 (2824, 2825, 2826)	3427
TICTOC Worklist # 26746 (2827, 2828, 2829)	3455
TICTOC Worklist # 26747 (2940, 2965)	3486
TICTOC Worklist # 26749 (2589, 2590, 2591)	3514
TICTOC Worklist # 26750 (2592, 2593, 2594)	3548
TICTOC Worklist # 26751 (2711, 2840)	3583
TICTOC Worklist # 26752 (2841, 2910, 2911)	3608
TICTOC Worklist # 26753 (2912, 2913)	3643
TICTOC Worklist # 26755 (2982)	3674
TICTOC Worklist # 26846 (3010, 3011)	3690
TICTOC Worklist # 26855 (3012, 3013)	3715
TICTOC Worklist # 27218 (2966)	3740
TICTOC Worklist # 27371 (2711, 2840)	3756
TICTOC Worklist # 27372 (2737, 2914)	3783
TICTOC Worklist # 27393 (2980, 2981, 2983)	3816
TICTOC Worklist # 27394 (2969)	3856
TICTOC Worklist # 27629 (3385)	3872
TICTOC Worklist # 27630 (3386, 3391, 3396)	3891
TICTOC Worklist # 28061 (2391)	3928
TICTOC Worklist # 28635 (2967, 2968)	3944
Radiochemical Analysis	3966
Total Alpha Analysis (AT)	
AT Worklist # 26718 (2806, 2807, 2808, 2809, 2810)	3968
AT Worklist # 26719 (2811, 2812, 2813, 2938, 2955)	3982
AT Worklist # 26720 (2956, 2957, 2958, 2959)	3999
AT Worklist # 26838 (3002, 3003, 3004, 3005)	4011
AT Worklist # 27092 (2598, 2599, 2600)	4024
AT Worklist # 27093 (2845, 2918)	4035
AT Worklist # 27103 (2601, 2602, 2603)	4046
AT Worklist # 27132 (2712, 2739, 2844)	4057
AT Worklist # 27133 (2920, 2921, 2922)	4069
AT Worklist # 27134 (2986, 2987)	4081
AT Worklist # 27135 (2988, 2989)	4090
AT Worklist # 27623 (3383)	4098
AT Worklist # 27899 (3393, 3398, 3388)	4105
AT Worklist # 28057 (2391)	4118

Sec 11
of 17Sec 12
of 17Sec 13
of 17

TABLE OF CONTENTS (Continued)

Sec 14
of 17

Gamma Energy Analysis (GEA)

GEA Worklist # 26735 (2814, 2815, 2816, 2817)	4124
GEA Worklist # 26737 (2818, 2819, 2820, 2821, 2939)	4167
GEA Worklist # 26738 (2960, 2961, 2962, 2963, 2964)	4218
GEA Worklist # 26841 (3006, 3007, 3008, 3009)	4269
GEA Worklist # 26978 (2598, 2599, 2600)	4312
GEA Worklist # 26979 (2601, 2602, 2603)	4347
GEA Worklist # 26980 (2845, 2918, 2919)	4381
GEA Worklist # 27141 (2712, 2739, 2844)	4416
GEA Worklist # 27142 (2920, 2921, 2922)	4451
GEA Worklist # 27143 (2986, 2987)	4486
GEA Worklist # 27144 (2988, 2989)	4512
GEA Worklist # 27168 (2598)	4538
GEA Worklist # 27278 (2986)	4556
GEA Worklist # 27288 (2598)	4574
GEA Worklist # 27626 (3384)	4592
GEA Worklist # 27865 (3393, 3398, 3388)	4610
GEA Worklist # 28056 (2391)	4645

Sec 15
of 17

Strontium 90 Analysis (Sr90)

Sr90 Worklist # 26725 (2818, 2819, 2820, 2821, 2939)	4663
Sr90 Worklist # 26968 (2598, 2599, 2600)	4678
Sr90 Worklist # 26969 (2601, 2602, 2603)	4684
Sr90 Worklist # 26970 (2845, 2918, 2919)	4694
Sr90 Worklist # 27091 (2598, 2599, 2600)	4704
Sr90 Worklist # 27136 (2712, 2739, 2844)	4714
Sr90 Worklist # 27137 (2920, 2921, 2922)	4724
Sr90 Worklist # 27138 (2986, 2987)	4734
Sr90 Worklist # 27139 (2988, 2989)	4742
Sr90 Worklist # 27191 (2919)	4750
Sr90 Worklist # 27304 (3006, 3007, 3008, 3009)	4756
Sr90 Worklist # 27411 (2814, 2815, 2816, 2817)	4768
Sr90 Worklist # 27546 (2960, 2961, 2962, 2963, 2964)	4780
Sr90 Worklist # 27669 (3384)	4795
Sr90 Worklist # 27743 (2988)	4801
Sr90 Worklist # 27864 (3393, 3398, 3388)	4807
Sr90 Worklist # 28060 (2391)	4817

Sec 16
of 17

Plutonium 239 Analysis (Pu239)

Pu239 Worklist # 26721 (2814, 2815, 2816, 2817)	4823
Pu239 Worklist # 26722 (2818, 2819, 2820, 2821, 2939)	4866
Pu239 Worklist # 26723 (2960, 2961, 2962, 2963, 2964)	4918
Pu239 Worklist # 26839 (3006, 3007, 3008, 3009)	4970
Pu239 Worklist # 27632 (3384)	5012
Pu239 Worklist # 28059 (2391)	5030

Sec 17
of 17

TABLE OF CONTENTS (Continued)

Sec 17
of 17
Continued

Opportunistic Analytes 5048

This document consists of pages 1 to 5113. Pages ii, 2, 32, 76, 80, 162, 202, 240, 242, 293, 326, 349, 3967 and 5049 were intentionally left blank.

This document contains pages 78.1, 940.1 and 1955.1.

TRADEMARKS:

Perkin-Elmer is a Registered Trademark of Research and Manufacturing Company, Inc.

Dionex is a Trademark of the Dionex Corporation.

LABCORE Completed Worklist Report for Worklist# 26721

Analyst: crj

Instrument: AB14

Book#

Method: Rev/Mod

Worklist Comment: BY-105 FOR @PU23901 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@PU23901 PU23901 LIQUID		1.12E+04	1.26E+4	112.500 % Recovery	
1 STD	0	@PU23901 PU23901E LIQUID		1.00	2.05E+00	2.050 % Ct Error	
1 STD	0	@PU23901 PU23901T LIQUID		100	7.85E+01	78.500 % Recovery	
2 BLNK	0	@PU23901 PU23901 LIQUID		1	<3.19E-6	uCi/mL	
2 BLNK	0	@PU23901 PU23901T LIQUID		100	8.49E+01	84.900 % Recovery	
2 BLNK	0	@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000 uCi/mL	
3 SAMPLE	S98TU02332	0 @PU23901 PU23901 LIQUID		N/A	3.20E-06	3.20E-006 uCi/mL	
3 SAMPLE	S98TU02332	0 @PU23901 PU23901E LIQUID		N/A	5.62E+01	% Recovery	
3 SAMPLE	S98TU02332	0 @PU23901 PU23901T LIQUID		N/A	1.00E+02	% Ct. Error	
4 DUP	S98TU02332	0 @PU23901 PU23901 LIQUID		<3.20E-6	<3.07E-6	RPD	
4 DUP	S98TU02332	0 @PU23901 PU23901E LIQUID		100	7.63E+01	76.300 % Recovery	
4 DUP	S98TU02332	0 @PU23901 PU23901T LIQUID		1.00	1.00E+02	100.000 % Ct. Error	
5 SAMPLE	S98TU02814	0 @PU23901 PU23901 LIQUID		N/A	3.67E-05	3.67E-005 uCi/mL	
5 SAMPLE	S98TU02814	0 @PU23901 PU23901T LIQUID		N/A	8.48E+01	% Recovery	
5 SAMPLE	S98TU02814	0 @PU23901 PU23901E LIQUID		N/A	1.22E+01	% Ct. Error	
6 DUP	S98TU02814	0 @PU23901 PU23901 LIQUID		<3.67E-5	<3.58E-5	RPD	
6 DUP	S98TU02814	0 @PU23901 PU23901T LIQUID		100	8.58E+01	85.800 % Recovery	
6 DUP	S98TU02814	0 @PU23901 PU23901E LIQUID		1.00	1.25E+01	12.500 % Ct. Error	
7 SAMPLE	S98TU02815	0 @PU23901 PU23901 LIQUID		N/A	4.22E-05	4.22E-005 uCi/mL	
7 SAMPLE	S98TU02815	0 @PU23901 PU23901T LIQUID		N/A	7.63E+01	% Recovery	
7 SAMPLE	S98TU02815	0 @PU23901 PU23901E LIQUID		N/A	1.00E+02	% Ct. Error	
8 DUP	S98TU02815	0 @PU23901 PU23901 LIQUID		<4.22E-5	<3.87E-5	RPD	
8 DUP	S98TU02815	0 @PU23901 PU23901T LIQUID		100	8.29E+01	82.900 % Recovery	
8 DUP	S98TU02815	0 @PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000 % Ct. Error	
9 SAMPLE	S98TU02816	0 @PU23901 PU23901 LIQUID		N/A	3.75E-05	3.75E-005 uCi/mL	
9 SAMPLE	S98TU02816	0 @PU23901 PU23901T LIQUID		N/A	8.30E+01	% Recovery	
9 SAMPLE	S98TU02816	0 @PU23901 PU23901E LIQUID		N/A	1.00E+02	% Ct. Error	
10 DUP	S98TU02816	0 @PU23901 PU23901 LIQUID		<3.75E-5	<4.14E-5	RPD	
10 DUP	S98TU02816	0 @PU23901 PU23901T LIQUID		100	7.85E+01	78.500 % Recovery	
10 DUP	S98TU02816	0 @PU23901 PU23901E LIQUID		1.00	1.28E+01	12.800 % Ct. Error	
11 SAMPLE	S98TU02817	0 @PU23901 PU23901 LIQUID		N/A	4.03E-05	4.03E-005 uCi/mL	
11 SAMPLE	S98TU02817	0 @PU23901 PU23901T LIQUID		N/A	7.94E+01	% Recovery	
11 SAMPLE	S98TU02817	0 @PU23901 PU23901E LIQUID		N/A	1.22E+01	% Ct. Error	
12 DUP	S98TU02817	0 @PU23901 PU23901 LIQUID		<4.03E-5	<4.14E-5	RPD	
12 DUP	S98TU02817	0 @PU23901 PU23901T LIQUID		100	7.79E+01	77.900 % Recovery	
12 DUP	S98TU02817	0 @PU23901 PU23901E LIQUID		1.00	1.12E+01	11.200 % Ct. Error	

Final page for worklist# 26721

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Completed Worklist Report for Worklist# 26721

Seq	Type	Sample#	RA	Test	Matrix	Actual	Found	DL or Yield	Unit

---

Analyst Signature	Date						Analyst Signature	Date	

<i>John Relyea</i>	<i>2 Dec 98</i>
Reviewer Signature	Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26721

Analyst: CRT Instrument: PU01 14 Book# 28B57

Method: LA-953-104 Rev/Mod B-D

Worklist Comment: BY-105 FOR @PU23901 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@PU23901	LIQUID		
2 BLNK			@PU23901	LIQUID		
~~3 SAMPLE~~	~~S98T002392 0~~		~~@PU23901~~	~~LIQUID~~	~~98000436~~	~~BY-105 (2)~~
~~Analytes Requested: PU23901 , PU23901E, PU23901T~~						
~~4 DUP~~	~~S98T002392 0~~		~~@PU23901~~	~~LIQUID~~		
5 SAMPLE	S98T002814 0		@PU23901	LIQUID	98000437	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T						
6 DUP	S98T002814 0		@PU23901	LIQUID		
7 SAMPLE	S98T002815 0		@PU23901	LIQUID	98000437	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T						
8 DUP	S98T002815 0		@PU23901	LIQUID		
9 SAMPLE	S98T002816 0		@PU23901	LIQUID	98000437	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T						
10 DUP	S98T002816 0		@PU23901	LIQUID		
11 SAMPLE	S98T002817 0		@PU23901	LIQUID	98000437	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T						
12 DUP	S98T002817 0		@PU23901	LIQUID		

Final page for worklist # 26721

Analyst Signature [Signature] Date 12/3/98

Analyst Signature [Signature] Date 12/4/98

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID

WORKBOOK PAGE: STD

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))				LIQUID	STD
Type	DATE COUNTED	DEC-03-98		Pu 236 AEA FRAC (C236)	0.358
STD	SAMPLE VOLUME in mL	SS	1.000	Pu 238 AEA FRAC (C238)	0.014
Work Use	SAMPLE DILUTION FACTOR	DF	1.000	Pu 239 AEA FRAC (C239)	0.547
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	3516
Std Code	DIGEST DILUTION FACTOR	DDF	1.000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43		BACKGROUND in cpm (Bkg)	0.030
Matrix	DETECTOR NUMBER	14		Pu 236 cpm	31.590
LIQUID	EFFICIENCY FACTOR	EFF	0.291	Pu 238 cpm	1.240
Batch Number	TRACER PREPARATION DATE	01/19/98		Pu 239 cpm	48.280
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000		AEA COUNT TIME	480
Re-run	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.810		Pu 239/240 μ Ci/L	1.2642E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000			
Sample Prep	STANDARD BOOK NO	28B57			
N/A	STANDARD VALUE in μ Ci/mL	1.120E-04			
Sample					
WL26721-STD					
Instrument Code	Decay Time = Date Counted - Tracer Preparation Date				
WB27807	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of ((-ln2 * Decay Time/1040.95))]				
Prepared By	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236 * 100 / Pu-236 Decay Corr'd Value * SPKV				
NEW	Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) / (1000mL/L) (DF) (DDF) / ((C236) (SS) (2220000 dpm/ μ Ci))				
Chemist	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)				
JFR	Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]				
Analyst	Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100				
CRJ					
Date Complete					
12/04/98	Pu 239/240 μ Ci/mL	1.26E-04	DETECTION LEVELS in μCi/mL Pu 239/240 8.84E-06		
Analysis Date	Relative Counting Error	= 2.0%			
12/03/98					
Analysis Time					
02:15 PM					
Sample ID					
BY-105 (2)	Pu 236 Tracer Recovery	= 78.5%			

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

STANDARD.WB1 REV 1.0

943128ML

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LI

BLNK

Type	DATE COUNTED	DEC-03-98	PU 236 AEA FRAC (C236)	0.917
BLNK	SAMPLE VOLUME in mL	SS	1.000 PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.000
26721	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1485
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cp (Bkg)	0.030
Matrix	DETECTOR NUMBER	14	PU 236 cpm	19.880
LIQUID	EFFICIENCY FACTOR	EFF	0.2910 PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.00	AEA COUNT TIME	480
Rerun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.81	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.189E-03
0	PU-238 TRACER VALUE (dpm/m)	0.00		

Sample Prep

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

WL26721-BLK

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of $\{-\ln 2 * \text{Decay Time}/1040.95\}$]

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WB27807

Pu 239/240 $\mu\text{Ci/L}$ = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

NEW

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Chemist

Relative Counting Error = Square Root of $\{[1 / (\text{Pu 236 cpm} * \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} * \text{min})]\} * 1.96 * 100$

JFR

Analyst

CRJ

Date Complete

12/04/98

Pu 239/240 $\mu\text{Ci/mL}$ < 3.19E-06

Analysis Date

Relative Counting Error = 100.0%

12/03/98

Analysis Time

NOTE: Pu 238 Result is a LESS THAN Value.

02:15 PM

Pu 238 $\mu\text{Ci/mL}$ < 3.19E-06

Sample Point

Relative Counting Error = 100.0%

BY-105 (2)

Pu 236 Tracer Recovery = 84.9%

DETECTION
LEVELS
in $\mu\text{Ci/mL}$

Pu 239/240

3.19E-06

Pu 238

3.19E-06

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

BLANK.WB1 REV 1.0

943128ML

4827

WORKBOOK PAGE: SAM5

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

SAMPLE

Type	DATE COUNTED	DEC-03-98	Pu 236 AEA FRAC (C236)	0.878
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	Pu 238 AEA FRAC (C238) 0.034
Wgt Pu List	SAMPLE DILUTION FACTOR	DF	11.000	Pu 239 AEA FRAC (C239) 0.014
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS 1549
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN) 30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.030
Matrix	DETECTOR NUMBER	14	Pu 236 cpm	34.600
LIQUID	EFFICIENCY FACTOR	EFF	0.291	Pu 238 cpm 1.340
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	0.550
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Refur	Pu-236 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.6687E-02
0	Pu-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002814

Instrument

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Date Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Pu 239/240 $\mu\text{Ci/mL}$	< 3.67E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
Relative Counting Error =	12.2%	
NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
Pu 238 $\mu\text{Ci/mL}$	< 3.67E-05	3.67E-05
Relative Counting Error =	7.9%	Pu 238
Pu 236 Tracer Recovery =	84.8%	3.67E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

WORKBOOK PAGE: DUP6

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

TYPE	DATE COUNTED	DEC-03-98	PU 236 AEA FRAC (C238)	0.890
DUP	SAMPLE VOLUME in mL	SS	1.000 PU 238 AEA FRAC (C238)	0.034
WORKLIST	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.014
26721	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1547
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.030
Batch Number	DETECTOR NUMBER	14	PU 238 cpm	34.270
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	1.290
	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.520
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Recon	PU-238 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.5751E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		
Sample Prep				
N/A				

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of ((-ln2 * Decay Time/1040.95)]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C238 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/04/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.58E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
Analysis Date	Relative Counting Error	= 12.5%	
12/03/98			Pu 239/240
Analysis time	NOTE: Pu 238 Result is a LESS THAN Value.		3.58E-05
02:15 PM	Pu 238 $\mu\text{Ci/mL}$	< 3.58E-05	Pu 238
Sample Point	Relative Counting Error	= 8.0%	3.58E-05
BY-105 (2)	Pu 236 Tracer Recovery	= 85.8%	

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

4829

WORKBOOK PAGE: SAM7

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

SAMPLE

Type	DATE COUNTED	DEC-03-98	PU 238 AEA FRAC (C236)	0.848
SAMPLE	SAMPLE VOLUME in mL	SS	1.000 PU 238 AEA FRAC (C238)	0.038
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.000
26721	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1444
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	146B43	BACKGROUND in cpm (Bkg)	0.030
Matrix	DETECTOR NUMBER	14	PU 236 cpm	21.270
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	0.950
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.009
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Room	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 $\mu\text{Ci/L}$ =	< 4.2191E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002815

Instrument Code

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Date Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Points

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Relative Counting Error = Square Root of { [1 / (Pu 236 cpm * min)] + [1 / (Pu 238 or 239/240 cpm * min)] } * 1.96 * 100

12/04/98	Pu 239/240 $\mu\text{Ci/mL}$	< 4.22E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
Analysis Date	Relative Counting Error	= 100.0%	
12/03/98			
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
02:15 PM	Pu 238 $\mu\text{Ci/mL}$	< 4.22E-05	4.22E-05
Sample Points	Relative Counting Error	= 9.4%	Pu 238
BY-105 (2)	Pu 236 Tracer Recovery	= 76.3%	4.22E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	<i>John Relyea</i>	Date:	7 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

4830

WORKBOOK PAGE: DUP8

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

DUP

Type	DATE COUNTED		DEC-03-98	PU 236 AEA FRAC (C236)	0.851
DUP	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.063
Work List	SAMPLE DILUTION FACTOR	DF	11.000	PU 239 AEA FRAC (C239)	0.000
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1563
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO		145B43	BACKGROUND in cpm (Bkg)	0.030
Matrix	DETECTOR NUMBER		14	PU 236 cpm	20.720
LIQUID	EFFICIENCY FACTOR	EFF	0.291	PU 238 cpm	1.520
Batch Number	TRACER PREPARATION DATE		01/19/98	PU 239 cpm	0.000
98005327	TRACER PREPARATION VALUE (dpm/mL)		2270.000	AEA COUNT TIME	480
ReRun	PU-236 DECAY CORR'D VALUE (dpm/mL)		1836.810	Pu 239/240 μ Ci/L =	< 3.8702E-02
0	PU-238 TRACER VALUE (dpm/mL)		0.000		

Sample Prep

N/A

Sample #

S98T002815

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Instrument Code

WB27807

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Prepared By

NEW

Pu 239/240 μ Ci/L = (C239) / (Pu 238 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Chemist

JFR

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C236] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Analyst

CRJ

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Date Complete

12/04/98

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Pu 239/240 μ Ci/mL < 3.87E-05

Relative Counting Error = 100.0%

DETECTION

LEVELS

in μ Ci/mL

Pu 239/240

3.87E-05

Pu 238

3.87E-05

Signature of Chemist

Signature of Analyst

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

Signature of Sample Point

4831

WORKBOOK PAGE: SAM9

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID / SO

	DATE COUNTED	DEC-03-98	PU 236 AEA FRAC (C236)	SAMPLE	0.877
TYPE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.000
Work Book	SAMPLE DILUTION FACTOR	DF	11.000	PU 239 AEA FRAC (C239)	0.000
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1518
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.030
Matrix	DETECTOR NUMBER	14	PU 236 cpm		32.010
LIQUID	EFFICIENCY FACTOR	EFF	0.291	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.000
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
Resin	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 μ Ci/L =	<	3.7522E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample #

S98T002816

Instrument Code

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Date Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C238) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Pu 239/240 μ Ci/mL	< 3.75E-05	DETECTION LEVELS
Relative Counting Error =	100.0%	in μ Ci/mL
NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
Pu 238 μ Ci/mL	< 3.75E-05	3.75E-05
Relative Counting Error =	100.0%	Pu 238
Pu 236 Tracer Recovery =	83.0%	3.75E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

4832

WORKBOOK PAGE: DUP10

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

Type	DATE COUNTED	DEC-03-98	Pu 236 AEA FRAC (C236)	0.840	
DUP	SAMPLE VOLUME in mL	SS	1.000	Pu 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	11.000	Pu 239 AEA FRAC (C239)	0.019
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1500
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.030
Matrix	DETECTOR NUMBER	14	Pu 236 cpm		21.680
LIQUID	EFFICIENCY FACTOR	EFF	0.291	Pu 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm		0.500
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
Regrun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 μ Ci/L =	<	4.1392E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample

S987002816

Instrument Code

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Date Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times$ [e to the power of $\{(-\ln 2 \times \text{Decay Time}/1040.95)\}$]Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times$ C236 $\times$ 100 / (Pu-236 Decay Corr'd Value $\times$ SPKV $\times$ EFF)Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]Pu 238 dpm = [(Total AT Counts / TC) - Bkg $\times$ 1/EFF $\times$ C238] - (Pu-238 Tracer Value $\times$ SPKV $\times$ Pu 236 Tracer Recovery / 100)Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]Relative Counting Error = Square Root of $\{[1 / (\text{Pu 236 cpm} \times \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

12/04/98	Pu 239/240 μ Ci/mL	< 4.14E-05	DETECTION LEVELS in μ Ci/mL
Analysis Date	Relative Counting Error	= 12.8%	
12/03/98			
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		
02:15 PM	Pu 238 μ Ci/mL	< 4.14E-05	
Sample Point	Relative Counting Error	= 100.0%	Pu 239/240
BY-105 (2)	Pu 236 Tracer Recovery	= 78.5%	4.14E-05
			Pu 238
			4.14E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128.M

4833

WORKBOOK PAGE: SAM11

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))				LIQUID /	SAMPLE	
TYPE	DATE COUNTED	DEC-03-98	PU 236 AEA FRAC	(C236)	0.853	
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC	(C238)	0.049
Work List	SAMPLE DILUTION FACTOR	DF	11.000	PU 239 AEA FRAC	(C239)	0.019
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1494	
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30	
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm	(Bkg)	0.030	
Matrix	DETECTOR NUMBER	14	PU 236 cpm		25.190	
LIQUID	EFFICIENCY FACTOR	EFF	0.291	PU 238 cpm	1.440	
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.550	
98005327	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480	
Return	PU-236 DECAY CORR'D VALUE (dpm/mL)	1836.810	Pu 239/240 μ Ci/L =	<	4.0301E-02	
0	PU-238 TRACER VALUE (dpm/mL)	0.000				

Sample Prep

N/A

Sample #

S98T002817

Instrument Code

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Date Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) * (Pu 236 Decay Corr'd Value) * (SPKV) * (1000mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/04/98	Pu 239/240 μ Ci/mL	< 4.03E-05	DETECTION LEVELS in μ Ci/mL
12/03/98	Relative Counting Error	= 12.2%	
02:15 PM	Pu 238 μ Ci/mL	5.23E-05	Pu 239/240
Sample Point	Relative Counting Error	= 7.7%	4.03E-05
BY-105 (2)	Pu 236 Tracer Recovery	= 79.4%	Pu 238
			4.03E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4834

WORKBOOK PAGE: DUP12

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID / SO

DUP

Type	DATE COUNTED		DEC-03-98	PU 236 AEA FRAC (C236)	0.846
DUP	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	11.000	PU 239 AEA FRAC (C239)	0.016
26721	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1478
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BQK NO		145B43	BACKGROUND in cpm (Bkg)	0.030
Matrix	DETECTOR NUMBER		14	PU 236 cpm	33.240
LIQUID	EFFICIENCY FACTOR	EFF	0.291	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE		01/19/98	PU 239 cpm	0.660
98005327	TRACER PREPARATION VALUE (dpm/mL)		2270.000	AEA COUNT TIME	480
Result	PU-236 DECAY CORR'D VALUE (dpm/mL)		1836.810	Pu 239/240 μ Ci/L =	< 4.141E-02
0	PU-238 TRACER VALUE (dpm/mL)		0.000		

Sample Prep

N/A

Sample #

S98T002817

Instrument Code

WB27807

Prepared By

NEW

Chemist

JFR

Analyst

CRJ

Data Complete

12/04/98

Analysis Date

12/03/98

Analysis Time

02:15 PM

Sample Point

BY-105 (2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Pu 239/240 μ Ci/mL < 4.14E-05

Relative Counting Error = 11.2%

DETECTION

LEVELS

in μ Ci/mL

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 238 μ Ci/mL < 4.14E-05

Relative Counting Error = 100.0%

Pu 236 Tracer Recovery = 77.9%

Pu 239/240

4.14E-05

Pu 238

4.14E-05

Analyst:	CRJ	Date:	04-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4835

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL26721-STD-PU
File ID: 1a1864.CNF

Counted on: 12/ 3/98 @21:25
Detector: AEA1
Geometry number: 1
Count time: 28806. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1303.1	1303.1	362.017	362.017	10.000	2.542	5.000	1.279
2	35.1	35.1	302.407	302.355	36.000	12.696	18.000	18.732
3	75.0	75.0	287.694	287.516	10.000	2.504	5.000	1.927
4?	20.8	20.8	270.418	270.073	12.000	3.340	6.000	1.432
5	2361.9	2361.9	229.578	229.578	10.000	2.685	5.000	1.965

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.358	5.755	5.756	-.0010.01	31.59	1.6	132.8	0.598E-04
2	Am241	0.014	5.479	5.481	-.0020.06	1.24	8.4	5.4	0.245E-05
	Pu238		5.487	5.481	0.006			7.1	0.320E-05
3	Th228	0.017	5.400	5.413	-.0130.01	1.47	8.1	8.5	0.385E-05
4	????			5.333		0.59	14.0		
5	Pu239	0.547	5.147	5.147	0.0000.01	48.28	1.3	198.9	0.896E-04
	Pu240		5.144	5.147	-.003			198.9	0.896E-04

Totals: 0.936 <--valid peaks only--> 82.58

DETECTOR CALIBRATION

Energy(MEV) = 4.090 + (0.0046)*Channel
Energy range (MeV): 4.090 TO 6.446
Efficiency = 0.2427 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	42366.0	100.000
Smoothed	42366.0	100.000
Composite fit	39925.7	94.240
Residuals	2440.3	5.760

Spectrum 1a1864.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 14584.0

.
.
..5..
.....5.
.....5
.....5..
.
4
4
4
4
3
3
2
2
2
.
.
..1
.....1.
.....1
.....1
.....1
..

Raw Data Dump for AEA Spectrum: 1a1864.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	1.	3.	0.	1.	1.	1.	1.	2.	2.
31	2.	0.	2.	1.	0.	0.	3.	2.	0.
41	2.	0.	0.	3.	3.	2.	1.	1.	0.
51	0.	2.	1.	4.	2.	0.	3.	3.	0.
61	3.	0.	1.	3.	0.	2.	2.	1.	2.
71	1.	0.	1.	3.	0.	1.	1.	1.	2.
81	2.	3.	2.	1.	3.	3.	2.	0.	3.
91	2.	2.	6.	2.	1.	7.	2.	4.	5.
101	3.	1.	2.	4.	3.	4.	3.	1.	3.
111	3.	5.	2.	3.	6.	2.	4.	3.	5.
121	7.	2.	5.	4.	5.	7.	6.	4.	5.
131	5.	5.	4.	6.	10.	4.	4.	8.	9.
141	6.	5.	4.	9.	5.	7.	8.	5.	8.
151	10.	13.	4.	7.	5.	2.	6.	5.	6.
161	8.	14.	13.	16.	6.	9.	12.	11.	11.
171	12.	13.	18.	19.	15.	11.	13.	7.	14.
181	18.	19.	20.	17.	18.	19.	16.	22.	19.
191	29.	24.	26.	38.	33.	29.	21.	32.	34.
201	48.	37.	54.	56.	46.	54.	64.	70.	88.
211	106.	105.	137.	172.	214.	234.	319.	427.	518.
221	552.	589.	648.	841.	1096.	1409.	1861.	2378.	2884.
231	2640.	1738.	787.	212.	34.	7.	4.	2.	4.
241	7.	7.	3.	8.	1.	3.	0.	8.	4.
251	7.	3.	6.	6.	9.	3.	5.	1.	9.
261	6.	6.	9.	11.	9.	10.	16.	19.	24.
271	31.	18.	17.	12.	21.	11.	6.	12.	15.
281	20.	30.	45.	38.	69.	72.	96.	105.	81.
291	41.	19.	6.	14.	21.	20.	18.	19.	21.
301	44.	24.	44.	27.	35.	19.	17.	16.	12.
311	15.	7.	10.	10.	12.	8.	15.	11.	8.
321	13.	17.	14.	19.	12.	8.	28.	21.	22.
331	17.	37.	24.	37.	32.	38.	43.	36.	46.
341	59.	74.	77.	89.	128.	138.	146.	196.	292.
351	474.	593.	613.	584.	570.	470.	565.	696.	801.
361	1470.	1625.	1644.	1171.	594.	173.	35.	4.	2.
371	2.	1.	1.	0.	1.	0.	1.	0.	0.
381	0.	1.	0.	1.	2.	0.	0.	0.	0.
391	1.	2.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	1.	1.	1.	1.	0.	0.
411	0.	2.	1.	1.	0.	0.	0.	2.	1.
421	0.	0.	2.	0.	0.	0.	3.	0.	0.
431	1.	0.	2.	1.	1.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	1.	0.	0.	1.
451	1.	0.	0.	0.	0.	0.	1.	0.	1.
461	1.	1.	0.	2.	3.	3.	4.	5.	2.
471	3.	9.	15.	9.	9.	6.	8.	1.	3.
481	2.	1.	1.	0.	0.	0.	1.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL26721-BLK-PU
 File ID: 2a2062.CNF

Counted on: 12/ 3/98 @21:25
 Detector: AEA2
 Geometry number: 1
 Count time: 28805. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	749.2	749.2	362.529	362.529	10.000	2.819	5.000	1.259
2	44.1	44.1	287.945	287.941	10.000	3.485	5.000	2.691
3	12.9	12.9	270.338	270.072	8.000	2.153	4.000	1.453

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count	%err	Activity
		Exp.	Obs.	Diff.	FWHM Rate c/m	@95	d/m uCi/ea
1	Pu236	0.917	5.755	5.753	0.0020.01	19.88	2.0 86.3 0.389E-04
2	Th228	0.044	5.400	5.410	-0.0100.02	0.96	9.2 5.7 0.258E-05
3		0.012		5.328	0.01	0.26	20.7 1.1 0.492E-06

Totals: 0.973 <--valid peaks only--> 21.09

DETECTOR CALIBRATION

Energy(MEV) = 4.085 + (0.0046)*Channel
 Energy range (MeV): 4.085 TO 6.441
 Efficiency = 0.2351 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	10406.0	100.000
Smoothed	10406.0	100.000
Composite fit	10125.7	97.307
Residuals	280.3	2.693

Analyzed by: _____

VR

Spectrum 2a2062.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 4199.1

```

3
3
3
..2
.2

1
.1
..... 1
.....1..
..... 1
.....1
..1...

```

Raw Data Dump for AEA Spectrum: 2a2062.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	1.	0.	0.	0.	0.	0.	0.	0.	1.	0.
31	1.	0.	1.	0.	0.	0.	0.	0.	0.	0.
41	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
61	0.	0.	0.	1.	0.	0.	0.	1.	0.	0.
71	1.	0.	0.	1.	0.	0.	0.	1.	0.	0.
81	0.	1.	0.	1.	0.	0.	1.	1.	0.	0.
91	0.	1.	0.	0.	0.	0.	0.	0.	0.	1.
101	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
111	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
121	0.	2.	0.	0.	0.	0.	1.	0.	1.	0.
131	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
141	1.	0.	0.	0.	0.	1.	0.	1.	0.	0.
151	1.	0.	1.	1.	0.	0.	0.	0.	0.	0.
161	0.	1.	0.	0.	0.	1.	0.	1.	0.	1.
171	1.	0.	0.	1.	0.	0.	1.	1.	0.	1.
181	0.	0.	1.	0.	0.	2.	0.	2.	1.	2.
191	2.	0.	0.	1.	2.	0.	1.	1.	4.	0.
201	1.	0.	2.	1.	0.	0.	0.	0.	2.	3.
211	1.	1.	2.	1.	1.	2.	1.	4.	1.	2.
221	0.	4.	3.	0.	2.	2.	3.	1.	4.	3.
231	0.	2.	1.	3.	1.	0.	3.	0.	1.	2.
241	2.	2.	1.	3.	1.	0.	2.	2.	2.	1.
251	0.	0.	1.	3.	2.	3.	3.	4.	2.	3.
261	2.	5.	3.	3.	6.	7.	7.	12.	18.	10.
271	24.	11.	8.	7.	1.	5.	9.	7.	9.	9.
281	15.	15.	12.	29.	44.	38.	50.	56.	50.	38.
291	26.	12.	2.	2.	6.	5.	5.	3.	4.	3.
301	5.	12.	13.	11.	8.	16.	6.	4.	5.	6.
311	4.	5.	4.	5.	5.	4.	3.	6.	8.	10.
321	13.	11.	6.	12.	11.	14.	9.	11.	14.	12.
331	9.	17.	10.	14.	15.	28.	24.	30.	30.	35.
341	34.	44.	56.	76.	66.	83.	100.	125.	163.	227.
351	292.	333.	394.	422.	350.	310.	312.	357.	504.	573.
361	741.	894.	943.	821.	575.	239.	81.	14.	2.	0.
371	0.	2.	1.	0.	0.	1.	1.	0.	0.	0.
381	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	1.	0.	0.	0.	0.	0.	1.	0.	0.
411	1.	0.	0.	1.	0.	0.	0.	1.	1.	1.
421	0.	0.	0.	0.	1.	0.	0.	0.	0.	1.
431	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
441	0.	1.	0.	0.	0.	0.	0.	1.	0.	1.
451	0.	1.	0.	0.	0.	0.	0.	1.	0.	2.
461	1.	0.	1.	0.	1.	1.	0.	3.	1.	1.
471	0.	2.	2.	12.	6.	7.	6.	3.	7.	1.
481	0.	2.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002814-SAM-P
File ID: 5a5284.CNF

Counted on: 12/ 3/98 @21:27
Detector: AEA5
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1303.8	1303.8	363.720	363.720	10.000	2.599	5.000	1.130
2	33.1	33.1	303.766	303.639	14.000	10.903	7.000	4.166
3	67.2	67.2	288.963	288.801	8.000	1.951	4.000	1.511
4	19.2	19.2	271.869	271.477	8.000	3.411	4.000	1.618
5	16.3	16.3	230.384	230.364	14.000	6.202	7.000	2.520

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.878	5.755	5.764	-0.0090.01	34.60	1.5	147.2	0.663E-04
	Cm243		5.779	5.764	0.015			197.7	0.890E-04
2	Am241	0.034	5.479	5.487	-0.0080.05	1.34	8.3	5.9	0.267E-05
	Pu238		5.487	5.487	0.000			7.7	0.349E-05
3	Th228	0.031	5.400	5.419	-0.0190.01	1.22	9.4	7.2	0.324E-05
4		0.013		5.339	0.02	0.51	16.2	2.1	0.965E-06
5	Pu239	0.014	5.147	5.150	-0.0030.03	0.55	12.2	2.3	0.103E-05
	Pu240		5.144	5.150	-0.006			2.3	0.103E-05

Totals:			0.970	<--valid peaks only-->		38.23			

DETECTOR CALIBRATION

Energy(MEV) = 4.090 + (0.0046)*Channel
Energy range (MeV): 4.090 TO 6.446
Efficiency = 0.2398 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18910.0	100.000
Smoothed	18909.7	99.998
Composite fit	18349.3	97.035
Residuals	560.7	2.965

Spectrum 5a5284.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

7874.6

5
5
5
54
4
423
23.
.22
2

..

.1.

.....1

.....1

.....1

..... 1

....1...

Raw Data Dump for AEA Spectrum: 5a5284.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	1.	1.	0.	1.	0.
31	0.	0.	0.	0.	0.	0.	0.	0.	1.
41	1.	0.	2.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	0.	0.
61	1.	0.	1.	1.	0.	1.	0.	0.	0.
71	0.	0.	0.	1.	1.	0.	1.	1.	0.
81	1.	0.	0.	1.	0.	0.	0.	0.	1.
91	0.	1.	1.	0.	2.	0.	0.	1.	0.
101	1.	1.	0.	0.	0.	1.	2.	1.	0.
111	0.	1.	0.	2.	0.	1.	0.	2.	0.
121	1.	1.	0.	0.	0.	0.	1.	0.	0.
131	0.	0.	2.	1.	1.	1.	0.	1.	0.
141	1.	1.	0.	1.	0.	1.	1.	1.	0.
151	0.	0.	0.	0.	1.	3.	0.	0.	1.
161	1.	1.	0.	1.	0.	1.	1.	1.	0.
171	0.	1.	2.	1.	1.	1.	0.	0.	1.
181	1.	0.	2.	1.	0.	2.	0.	2.	1.
191	4.	2.	0.	0.	1.	0.	5.	0.	0.
201	4.	3.	0.	2.	2.	1.	5.	5.	2.
211	0.	2.	3.	5.	5.	1.	4.	2.	9.
221	7.	6.	4.	9.	7.	17.	14.	18.	17.
231	14.	21.	15.	7.	6.	4.	2.	5.	4.
241	3.	4.	2.	4.	6.	2.	4.	3.	4.
251	7.	7.	7.	7.	8.	3.	5.	3.	6.
261	6.	4.	7.	12.	9.	8.	14.	20.	18.
271	24.	26.	29.	20.	10.	9.	18.	13.	11.
281	19.	22.	23.	29.	36.	68.	64.	84.	98.
291	64.	33.	14.	13.	7.	15.	11.	18.	16.
301	16.	19.	30.	22.	24.	19.	17.	18.	18.
311	19.	16.	15.	15.	14.	21.	23.	15.	23.
321	22.	26.	34.	27.	30.	30.	40.	43.	33.
331	41.	53.	44.	46.	54.	63.	56.	73.	67.
341	93.	103.	106.	122.	123.	170.	184.	205.	224.
351	391.	421.	585.	664.	651.	651.	573.	582.	645.
361	986.	1181.	1462.	1648.	1519.	1073.	446.	102.	21.
371	0.	1.	2.	2.	2.	0.	0.	1.	1.
381	0.	0.	0.	0.	0.	0.	0.	1.	3.
391	2.	0.	0.	0.	0.	1.	0.	0.	0.
401	0.	0.	0.	1.	0.	0.	0.	0.	1.
411	0.	1.	2.	1.	0.	1.	1.	0.	1.
421	0.	0.	3.	1.	2.	1.	1.	0.	1.
431	1.	0.	1.	1.	0.	1.	0.	1.	1.
441	1.	1.	1.	0.	0.	0.	0.	1.	1.
451	1.	0.	0.	0.	1.	0.	1.	0.	2.
461	0.	1.	1.	3.	2.	1.	0.	2.	1.
471	2.	3.	6.	4.	7.	9.	12.	15.	9.
481	8.	3.	1.	0.	1.	1.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	1.
511	1.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002814-DUP-P
File ID: 6a6289.CNF

Counted on: 12/ 3/98 @21:27
Detector: AEA6
Geometry number: 1
Count time: 28807. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1551.8	1551.8	360.886	360.886	10.000	2.294	5.000	1.310
2	18.4	18.4	302.534	302.504	36.000	48.715	18.000	21.454
3	63.9	63.9	286.303	286.198	10.000	1.742	5.000	1.399
4	22.1	22.1	268.963	268.736	8.000	1.969	4.000	1.503
5	20.3	20.3	228.239	228.237	12.000	3.763	6.000	2.048

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err	d/m	Activity	
			Exp.	Obs.	Diff.	FWHM	c/m @95		uCi/ea	
1	Pu236	0.890	5.755	5.753	0.0020	0.01	34.27	1.5	153.5 0.691E-04	
2	Am241	0.034	5.479	5.484	-0.0050	0.22	1.29	11.5	6.0 0.272E-05	
	Pu238		5.487	5.484	0.003				7.9 0.355E-05	
3	Th228	0.029	5.400	5.409	-0.0090	0.01	1.11	9.8	6.9 0.309E-05	
4		0.011		5.329	0.01		0.41	16.5	1.8 0.802E-06	
5	Pu239	0.014	5.147	5.143	0.0040	0.02	0.52	12.4	2.3 0.103E-05	
	Pu240		5.144	5.143	0.001				2.3 0.103E-05	
Totals:			0.977	<--valid peaks only-->			37.60			

DETECTOR CALIBRATION

Energy (MeV) = 4.093 + (0.0046)*Channel
Energy range (MeV): 4.093 TO 6.448
Efficiency = 0.2278 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18482.0	100.000
Smoothed	18481.9	100.000
Composite fit	18052.5	97.676
Residuals	429.5	2.324

4845 Analyzed by: _____

Spectrum 6a6289.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 6635.7

5
5
54
4
3
2.3.
3
2
2
2
2
1......1
.....1.....
.....
.....1.
.....

Raw Data Dump for AEA Spectrum: 6a6289.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	0.	0.	0.	0.	1.	0.	1.	0.	0.	0.
31	0.	1.	0.	1.	1.	1.	0.	0.	0.	0.
41	0.	0.	0.	0.	0.	1.	1.	0.	0.	1.
51	0.	1.	0.	0.	0.	0.	0.	0.	1.	1.
61	0.	0.	1.	0.	0.	1.	0.	1.	0.	2.
71	0.	0.	0.	0.	1.	1.	0.	2.	0.	1.
81	0.	1.	1.	0.	0.	1.	0.	0.	0.	0.
91	0.	1.	0.	0.	1.	0.	1.	0.	0.	0.
101	0.	0.	0.	1.	0.	1.	0.	0.	0.	1.
111	1.	0.	0.	0.	3.	0.	1.	0.	1.	0.
121	1.	0.	1.	0.	0.	0.	0.	0.	0.	1.
131	1.	0.	0.	2.	0.	0.	0.	1.	1.	1.
141	1.	0.	0.	0.	1.	1.	1.	1.	1.	2.
151	0.	1.	1.	0.	1.	1.	3.	0.	1.	0.
161	1.	2.	1.	0.	0.	2.	2.	0.	2.	1.
171	2.	0.	0.	1.	0.	1.	2.	2.	1.	1.
181	0.	0.	0.	4.	1.	0.	3.	1.	0.	3.
191	1.	2.	1.	0.	1.	0.	2.	2.	1.	2.
201	0.	4.	1.	3.	2.	1.	2.	0.	1.	4.
211	3.	3.	1.	3.	5.	8.	7.	2.	3.	8.
221	9.	10.	6.	13.	17.	17.	18.	25.	28.	18.
231	10.	8.	4.	4.	1.	2.	3.	3.	3.	3.
241	2.	2.	7.	3.	1.	3.	0.	3.	3.	6.
251	3.	5.	3.	1.	4.	3.	9.	4.	6.	3.
261	4.	6.	5.	10.	13.	16.	22.	22.	37.	29.
271	17.	8.	13.	9.	5.	12.	9.	19.	18.	22.
281	34.	35.	46.	56.	75.	96.	86.	83.	34.	10.
291	9.	4.	7.	9.	12.	8.	9.	15.	6.	12.
301	20.	12.	23.	17.	12.	8.	10.	11.	6.	10.
311	6.	10.	10.	6.	5.	8.	8.	5.	12.	12.
321	12.	15.	12.	19.	9.	15.	20.	19.	25.	17.
331	28.	25.	20.	37.	28.	49.	31.	40.	43.	56.
341	69.	72.	87.	130.	140.	164.	191.	251.	388.	542.
351	656.	716.	635.	523.	471.	567.	732.	995.	1323.	1719.
361	2082.	1875.	1245.	525.	132.	15.	2.	0.	0.	2.
371	0.	1.	1.	1.	0.	0.	0.	1.	0.	0.
381	0.	1.	0.	0.	1.	0.	0.	0.	0.	1.
391	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	1.	1.	0.	0.	1.
411	1.	0.	0.	0.	1.	0.	1.	1.	0.	1.
421	0.	1.	1.	0.	2.	0.	1.	1.	2.	0.
431	1.	2.	1.	0.	0.	1.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	1.	1.	0.	0.
451	0.	2.	0.	0.	0.	0.	2.	0.	1.	1.
461	0.	0.	0.	1.	3.	0.	1.	0.	0.	5.
471	9.	2.	5.	11.	10.	5.	10.	5.	1.	5.
481	1.	1.	0.	1.	1.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	1.	0.	0.	1.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002815-SAM-P
File ID: 7a7317.CNF

Counted on: 12/ 3/98 @21:27
Detector: AEA7
Geometry number: 1
Count time: 28806. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	856.4	856.4	363.646	363.646	8.000	2.248	4.000	1.040
2	15.3	15.3	307.181	307.181	78.000	1.000	39.000	0.100
3	11.8	11.8	304.290	303.050	64.000	1.000	32.000	0.100
4?	22.2	22.2	288.676	288.542	6.000	0.922	3.000	1.529
5?	6.0	6.0	270.527	270.318	6.000	0.597	3.000	1.286
6?	7.1	7.1	230.378	232.042	14.000	0.858	7.000	0.501

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Exp.	Obs.	Diff.	FWHM	Count Rate	%err	d/m	Activity
1	Pu236	0.848	5.755	5.764	-0.009	0.01	21.27	1.9	90.0	0.405E-04
	Cm243		5.779	5.764	0.015				120.8	0.544E-04
~~2~~	~~Pu238~~	~~0.049~~	~~5.487~~	~~5.505~~	~~-0.018~~	~~0.00~~	~~1.24~~	~~8.1~~	7.1	0.320E-05
3	Am241		5.479	5.505	-0.026				5.4	0.245E-05
	Pu238	0.038	5.487	5.486	0.001	0.00	0.95	13.9	5.5	0.246E-05
	Am241		5.479	5.486	-0.007				4.2	0.188E-05
4		????		5.419			0.24	49.9		
5		????		5.335			0.05	169.		
6		????		5.159			0.13	46.7		

Totals:		0.935	<--valid peaks only-->				23.45			

DETECTOR CALIBRATION

Energy (MEV) = 4.092 + (0.0046)*Channel
Energy range (MeV): 4.092 TO 6.447
Efficiency = 0.2413 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	12040.0	100.000
Smoothed	12040.0	100.000
Composite fit	11462.5	95.203
Residuals	577.5	4.797

Spectrum 7a7317.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

4927.9

```

6
6
6
3
3
3
5
3
4.
4.
3
3
.
.
.1..
....1.
.....1..
..... 1
..... 1
..1....

```

Raw Data Dump for AEA Spectrum: 7a7317.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	2.	0.	0.	0.	0.	0.	0.	0.	1.
31	0.	1.	0.	0.	0.	1.	0.	0.	0.
41	0.	0.	0.	0.	0.	0.	1.	0.	1.
51	0.	0.	0.	0.	2.	0.	0.	0.	0.
61	0.	1.	1.	0.	0.	0.	0.	1.	0.
71	0.	0.	1.	1.	0.	0.	0.	1.	0.
81	0.	0.	0.	0.	0.	0.	0.	1.	0.
91	0.	1.	1.	1.	0.	1.	3.	0.	2.
101	2.	1.	0.	2.	0.	0.	1.	0.	1.
111	0.	0.	0.	0.	0.	1.	0.	0.	0.
121	0.	0.	0.	2.	0.	0.	0.	2.	0.
131	0.	0.	1.	1.	0.	1.	1.	3.	0.
141	0.	1.	1.	1.	1.	0.	2.	1.	3.
151	1.	1.	2.	1.	2.	1.	0.	2.	0.
161	1.	2.	0.	2.	2.	0.	3.	0.	1.
171	2.	0.	2.	1.	2.	1.	1.	1.	0.
181	2.	1.	1.	0.	2.	0.	2.	1.	1.
191	2.	1.	3.	0.	2.	0.	2.	2.	2.
201	0.	1.	1.	3.	4.	4.	1.	2.	1.
211	2.	0.	1.	1.	2.	1.	5.	2.	1.
221	5.	1.	6.	4.	8.	6.	7.	8.	16.
231	6.	11.	12.	8.	3.	3.	2.	1.	6.
241	1.	3.	2.	1.	7.	6.	2.	4.	3.
251	6.	6.	4.	7.	8.	5.	2.	5.	4.
261	3.	4.	10.	5.	9.	10.	15.	11.	14.
271	23.	17.	8.	10.	6.	5.	10.	10.	14.
281	16.	17.	17.	19.	22.	31.	33.	43.	53.
291	21.	15.	11.	7.	11.	11.	8.	11.	14.
301	12.	12.	19.	16.	17.	11.	19.	13.	18.
311	12.	15.	7.	13.	15.	17.	13.	17.	24.
321	19.	25.	20.	29.	21.	26.	28.	23.	31.
331	38.	36.	33.	47.	34.	50.	43.	52.	61.
341	56.	61.	77.	97.	91.	92.	107.	150.	163.
351	233.	307.	396.	428.	427.	353.	324.	283.	382.
361	523.	753.	1012.	1147.	1017.	560.	190.	37.	6.
371	1.	0.	1.	0.	0.	1.	0.	0.	0.
381	1.	0.	0.	0.	0.	0.	0.	0.	1.
391	0.	0.	0.	0.	0.	1.	0.	1.	0.
401	0.	0.	0.	0.	0.	1.	0.	1.	1.
411	0.	0.	0.	0.	0.	0.	0.	1.	0.
421	0.	0.	0.	0.	0.	0.	1.	0.	0.
431	0.	0.	0.	0.	0.	0.	1.	1.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	0.	0.	0.	0.	0.	0.	0.
461	0.	1.	0.	0.	0.	0.	0.	2.	0.
471	1.	3.	2.	1.	5.	5.	9.	11.	3.
481	5.	4.	0.	1.	2.	2.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002815-DUP-P
File ID: 8a8295.CNF

Counted on: 12/ 3/98 @21:28
Detector: AEAB
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	734.5	734.5	362.712	362.712	8.000	2.444	4.000	0.948
2	18.9	18.9	304.126	304.010	70.000	1.000	35.000	0.100
3	27.1	27.1	288.240	288.125	10.000	1.856	5.000	1.549
4?	6.9	6.9	271.726	271.232	10.000	2.261	5.000	2.101
5?	8.0	8.0	230.753	230.587	10.000	3.052	5.000	1.889

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea	
			Exp.	Obs.	Diff.						
1	Pu236	0.851	5.755	5.761	-0.0060	0.01	20.72	2.0	91.1	0.410E-04	
	Cm243		5.779	5.761	0.018				122.3	0.551E-04	
2	Pu238	0.063	5.487	5.491	-0.0040	0.00	1.52	7.4	9.1	0.411E-05	
	Am241		5.479	5.491	-0.012				7.0	0.315E-05	
3	Th228	0.019	5.400	5.418	-0.0180	0.01	0.47	24.3	2.9	0.129E-05	
4	????			5.340			0.12	70.6			
5	????			5.153			0.18	32.3			
Totals:			0.933	<--valid peaks only-->			22.71				

DETECTOR CALIBRATION

Energy(MEV) = 4.093 + (0.0046)*Channel
Energy range (MeV): 4.093 TO 6.448
Efficiency = 0.2320 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	11688.0	100.000
Smoothed	11688.0	100.000
Composite fit	11049.2	94.534
Residuals	638.8	5.466

```

5
5
5
2
2
2
2
4
4
23.
3..
.2
2.
..
1...
...1..
.....1
.....1...
.....1
.....1
.....1....
.1....
```

Raw Data Dump for AEA Spectrum: 8a8295.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
31	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
41	0.	0.	1.	0.	0.	0.	0.	1.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
61	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
71	0.	1.	0.	0.	0.	0.	2.	1.	0.	0.
81	1.	0.	0.	0.	0.	0.	0.	0.	1.	0.
91	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
101	1.	0.	1.	1.	1.	0.	1.	1.	0.	0.
111	0.	0.	0.	2.	1.	0.	0.	0.	0.	0.
121	1.	0.	0.	1.	1.	1.	1.	0.	0.	0.
131	1.	0.	0.	0.	0.	0.	0.	0.	1.	0.
141	0.	0.	0.	0.	1.	2.	3.	0.	1.	0.
151	0.	0.	0.	2.	1.	0.	0.	0.	0.	2.
161	2.	0.	1.	0.	2.	1.	1.	0.	0.	0.
171	1.	0.	1.	1.	0.	2.	0.	0.	1.	0.
181	1.	2.	0.	3.	1.	0.	4.	0.	0.	0.
191	1.	1.	2.	0.	2.	2.	1.	1.	1.	1.
201	3.	0.	3.	2.	2.	2.	2.	1.	3.	2.
211	3.	3.	3.	1.	4.	3.	5.	5.	3.	3.
221	2.	4.	3.	6.	5.	9.	0.	14.	9.	14.
231	11.	11.	11.	6.	1.	4.	4.	6.	5.	1.
241	2.	1.	4.	4.	2.	5.	4.	6.	2.	5.
251	4.	5.	8.	3.	5.	4.	8.	13.	6.	2.
261	9.	6.	7.	7.	9.	12.	6.	12.	13.	16.
271	21.	11.	21.	11.	10.	16.	14.	8.	8.	11.
281	20.	23.	29.	17.	37.	24.	47.	45.	55.	32.
291	28.	15.	14.	10.	8.	12.	18.	11.	14.	13.
301	11.	15.	24.	24.	25.	18.	13.	11.	8.	15.
311	12.	11.	24.	19.	23.	18.	25.	21.	19.	20.
321	33.	33.	21.	33.	33.	33.	36.	29.	44.	51.
331	46.	44.	41.	65.	48.	68.	62.	55.	74.	84.
341	82.	80.	100.	115.	117.	136.	162.	174.	203.	255.
351	297.	355.	404.	403.	344.	322.	313.	368.	396.	501.
361	632.	841.	950.	876.	553.	209.	61.	9.	3.	2.
371	1.	0.	1.	1.	1.	1.	0.	1.	0.	0.
381	0.	0.	0.	0.	0.	2.	1.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
401	0.	2.	0.	1.	0.	0.	0.	0.	0.	0.
411	0.	2.	1.	0.	0.	0.	0.	0.	0.	0.
421	1.	0.	0.	0.	0.	0.	0.	2.	1.	0.
431	1.	0.	0.	1.	0.	0.	0.	1.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	0.	1.	2.	0.	0.	1.
461	0.	0.	0.	2.	0.	1.	1.	1.	3.	3.
471	0.	1.	1.	6.	6.	10.	4.	2.	6.	4.
481	3.	1.	1.	3.	0.	0.	0.	0.	0.	1.
491	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002816-SAM-P
File ID: 9a99208.CNF

Counted on: 12/ 3/98 @21:28
Detector: AEA9
Geometry number: 1
Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1037.6	1037.6	361.016	361.016	10.000	2.735	5.000	0.954
2	22.2	22.2	303.884	303.130	58.000	365.246	29.000	110.682
3	41.6	41.6	285.534	285.444	8.000	2.220	4.000	2.000
4?	7.5	7.5	268.747	268.323	6.000	1.153	3.000	1.697
5?	9.5	9.5	228.321	228.007	10.000	3.561	5.000	1.561

not a valid peak
JHR

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count		%err	Activity	
			Exp.	Obs. Diff.	FWHM	Rate	c/m @95	d/m	uCi/ea
1	Pu236	0.877	5.755	5.756	-.0010.01	32.01	1.6	135.0	0.608E-04
2	Am241	0.116	5.479	5.490	-.0111.68	4.22	12.8	18.5	0.835E-05
3	Pu238	5.487	5.490	-.003				24.2	0.109E-04
3	Th228	0.020	5.400	5.409	-.0090.01	0.74	19.3	4.3	0.194E-05
4	????			5.330		0.09	103.		
5	????			5.145		0.27	22.1		

Totals: 1.013 <--valid peaks only--> 36.97

DETECTOR CALIBRATION

Energy (MEV) = 4.096 + (0.0046)*Channel

Energy range (MeV): 4.096 TO 6.451

Efficiency = 0.2420 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17524.0	100.000
Smoothed	17524.0	100.000
Composite fit	17918.4	102.251
Residuals	-394.4	-2.251

Analyzed by:

VR

Spectrum 9a99208.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

5362.2

5
 5
 5
 2
 2
 2
 2
 4.
 3.
 .23..
 .2
 .2
 .2
 12.
 2.1.
 2.....1
 2..... 1
 2.....1....
 2..... 1
 2.....1
 2..

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	1.	0.	1.	0.	1.
21	0.	1.	0.	0.	0.	0.	0.	0.	1.	0.
31	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
41	0.	0.	0.	0.	0.	0.	2.	0.	1.	0.
51	0.	1.	0.	0.	0.	0.	0.	1.	1.	0.
61	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
71	0.	0.	0.	0.	0.	0.	2.	0.	1.	0.
81	1.	2.	0.	1.	0.	0.	0.	0.	0.	0.
91	0.	0.	0.	1.	1.	0.	1.	0.	0.	0.
101	1.	0.	0.	0.	1.	1.	0.	0.	1.	1.
111	2.	1.	0.	1.	1.	0.	1.	1.	0.	2.
121	0.	1.	0.	0.	1.	2.	0.	1.	1.	1.
131	1.	3.	0.	2.	0.	0.	3.	1.	1.	2.
141	0.	0.	0.	0.	0.	1.	3.	1.	2.	0.
151	2.	0.	0.	0.	0.	2.	0.	1.	2.	1.
161	1.	1.	2.	1.	1.	1.	0.	2.	1.	0.
171	1.	0.	2.	1.	1.	2.	0.	0.	0.	2.
181	2.	0.	0.	2.	1.	2.	1.	2.	0.	2.
191	1.	1.	3.	2.	2.	1.	1.	5.	3.	4.
201	1.	4.	2.	1.	2.	4.	2.	4.	2.	2.
211	8.	4.	2.	5.	3.	4.	7.	4.	7.	1.
221	5.	3.	5.	5.	12.	13.	14.	13.	16.	9.
231	14.	4.	8.	3.	3.	7.	2.	2.	7.	9.
241	7.	4.	6.	2.	4.	2.	4.	6.	4.	8.
251	14.	7.	8.	12.	5.	8.	7.	12.	12.	13.
261	10.	11.	22.	12.	17.	13.	29.	23.	26.	27.
271	18.	15.	18.	17.	13.	21.	25.	31.	33.	37.
281	35.	45.	54.	69.	62.	78.	65.	52.	29.	22.
291	20.	17.	22.	17.	13.	27.	15.	21.	28.	25.
301	20.	17.	18.	36.	20.	15.	30.	14.	16.	28.
311	25.	34.	27.	30.	39.	33.	30.	29.	46.	36.
321	32.	27.	39.	45.	50.	52.	65.	56.	61.	74.
331	63.	67.	90.	89.	82.	99.	108.	113.	112.	156.
341	125.	170.	202.	214.	236.	283.	284.	364.	437.	512.
351	544.	592.	622.	508.	505.	558.	604.	778.	908.	1108.
361	1244.	1268.	992.	536.	154.	34.	3.	0.	0.	0.
371	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
381	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
411	0.	1.	0.	0.	1.	1.	0.	0.	0.	1.
421	0.	0.	1.	1.	3.	0.	0.	0.	1.	0.
431	0.	1.	0.	0.	0.	1.	1.	0.	0.	0.
441	0.	0.	0.	0.	1.	0.	0.	1.	0.	0.
451	1.	1.	0.	0.	0.	0.	0.	0.	1.	2.
461	1.	0.	1.	1.	1.	3.	3.	1.	4.	0.
471	3.	5.	5.	9.	8.	5.	4.	4.	2.	1.
481	2.	1.	0.	0.	0.	0.	0.	0.	0.	0.
491	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T002816-DUP-P
 File ID: 10a1047.CNF

Counted on: 12/ 3/98 @21:29
 Detector: AEA10
 Geometry number: 1
 Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	624.1	624.1	361.746	361.746	10.000	3.004	5.000	0.900
2	41.8	41.8	287.509	287.465	12.000	10.034	6.000	2.635
3	9.4	9.4	229.984	229.901	12.000	8.655	6.000	1.789

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count	%err	Activity
		Exp.	Obs.	Diff.	FWHM	Rate c/m @95	d/m uCi/ea
1	Pu236	0.840	5.755	5.747	0.0080.01	21.68 1.9	95.7 0.431E-04
2	Th228	0.076	5.400	5.405	-0.0050.05	1.97 6.6	12.0 0.541E-05
3	Pu239	0.019	5.147	5.140	0.0070.04	0.50 13.6	2.2 0.971E-06
	Pu240		5.144	5.140	0.004		2.2 0.971E-06

Totals:		0.936	<--valid peaks only-->			24.15	

DETECTOR CALIBRATION

Energy(MEV) = 4.083 + (0.0046)*Channel
 Energy range (MeV): 4.083 TO 6.438
 Efficiency = 0.2312 CPM/DCM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	12389.0	100.000
Smoothed	12389.0	100.000
Composite fit	11593.3	93.577
Residuals	795.7	6.423

Analyzed by:

VR

Spectrum 10a1047.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

3224.9

3
3
3
3
3
3.
2

.2

..2

....2

....2

...

1...

.1...

....1...

.....1..

.....1

.....1..

.....1

.....1

.1...

1

Raw Data Dump for AEA Spectrum: 10a1047.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	1.	0.	1.
21	0.	0.	0.	0.	0.	2.	0.	0.	0.
31	0.	0.	0.	0.	0.	0.	0.	0.	0.
41	0.	1.	0.	0.	0.	0.	1.	1.	0.
51	0.	0.	3.	0.	0.	0.	2.	0.	2.
61	1.	0.	0.	1.	0.	0.	0.	2.	0.
71	0.	0.	0.	0.	0.	1.	0.	1.	0.
81	0.	1.	0.	1.	1.	1.	0.	0.	1.
91	0.	1.	1.	1.	1.	0.	0.	0.	1.
101	1.	0.	3.	2.	0.	1.	0.	2.	1.
111	0.	0.	1.	1.	1.	0.	1.	0.	1.
121	0.	0.	0.	0.	0.	1.	0.	2.	1.
131	1.	0.	1.	0.	1.	2.	0.	1.	3.
141	1.	1.	0.	1.	1.	2.	0.	3.	1.
151	2.	1.	1.	0.	2.	2.	0.	1.	2.
161	1.	1.	0.	3.	3.	1.	1.	1.	0.
171	1.	1.	1.	2.	1.	1.	3.	1.	2.
181	2.	1.	0.	1.	1.	1.	2.	6.	1.
191	1.	3.	3.	2.	2.	3.	3.	5.	2.
201	3.	5.	1.	1.	1.	2.	2.	3.	2.
211	4.	3.	7.	5.	3.	10.	4.	6.	6.
221	9.	8.	4.	6.	8.	6.	6.	9.	8.
231	11.	9.	4.	4.	5.	4.	4.	3.	7.
241	3.	4.	5.	9.	7.	7.	5.	4.	8.
251	7.	10.	7.	7.	9.	8.	5.	7.	6.
261	10.	12.	13.	13.	11.	15.	16.	22.	18.
271	15.	17.	14.	16.	15.	18.	18.	15.	11.
281	27.	28.	36.	31.	33.	47.	50.	51.	45.
291	26.	17.	18.	11.	18.	17.	13.	20.	19.
301	20.	21.	26.	25.	26.	24.	24.	27.	26.
311	28.	32.	27.	29.	21.	30.	29.	39.	37.
321	40.	42.	45.	51.	43.	54.	56.	73.	73.
331	77.	74.	86.	61.	83.	94.	83.	100.	90.
341	117.	109.	164.	144.	162.	176.	184.	235.	259.
351	325.	355.	389.	396.	367.	358.	342.	419.	509.
361	684.	753.	733.	525.	287.	117.	24.	3.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	1.
411	0.	0.	0.	0.	0.	0.	0.	0.	1.
421	0.	0.	0.	1.	1.	0.	1.	0.	0.
431	0.	1.	1.	0.	1.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	1.	0.	0.	1.	0.	0.	0.	1.	1.
461	0.	0.	0.	0.	1.	0.	0.	3.	2.
471	2.	3.	6.	5.	4.	3.	4.	6.	2.
481	4.	2.	1.	0.	0.	0.	0.	0.	1.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T002817-SAM-P
File ID: 11a1144.CNF

Counted on: 12/ 3/98 @21:29
Detector: AEA11
Geometry number: 1
Count time: 28806. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	823.5	823.5	364.211	364.211	10.000	2.659	5.000	0.932
2	17.9	17.9	308.554	308.087	62.000	1.000	31.000	0.100
3	40.5	40.5	289.763	289.659	8.000	2.305	4.000	1.787
4?	12.7	12.7	273.687	273.143	16.000	3.453	8.000	2.925
5	23.2	23.2	231.978	231.946	10.000	3.279	5.000	1.966

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Exp.	Centroid Obs.	Diff.	FWHM	Count Rate	%err c/m	@95	d/m	Activity uCi/ea
1	Pu236	0.853	5.755	5.750	0.0050	0.01	25.19	1.8		107.8	0.485E-04
2	Pu238	0.049	5.487	5.492	-.0050	0.00	1.44	7.8		8.4	0.377E-05
	Am241		5.479	5.492	-.013					6.4	0.289E-05
3	Th228	0.026	5.400	5.407	-.0070	0.01	0.77	15.9		4.6	0.205E-05
4	????			5.331			0.27	35.7			
5	Pu239	0.019	5.147	5.142	0.0050	0.02	0.55	14.4		2.3	0.103E-05
	Pu240		5.144	5.142	0.002					2.3	0.103E-05

Totals:		0.947	<--valid peaks only-->				27.95				

DETECTOR CALIBRATION

Energy(MEV) = 4.075 + (0.0046)*Channel
Energy range (MeV): 4.075 TO 6.430
Efficiency = 0.2385 CPM/CPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	14175.0	100.000
Smoothed	14175.3	100.002
Composite fit	13545.5	95.559
Residuals	629.5	4.441

Spectrum 11a1144.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

5167.3

5
5
5
2
2
2
4
4.
3.
23.
2
2.
..
.1.
...1..
.....1
.....1
.....1
.....1
.....1
.....1

Raw Data Dump for AEA Spectrum: 11a1144.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	1.	1.	1.	0.
21	0.	0.	0.	1.	0.	0.	1.	1.	0.	1.
31	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
41	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	1.	1.	0.	1.	2.	1.	1.
61	1.	0.	0.	0.	0.	1.	0.	0.	1.	0.
71	0.	0.	0.	2.	0.	0.	0.	0.	1.	0.
81	0.	0.	0.	0.	2.	0.	1.	1.	0.	0.
91	0.	1.	1.	1.	0.	1.	2.	0.	1.	0.
101	0.	0.	0.	0.	0.	1.	0.	1.	0.	1.
111	2.	0.	0.	0.	1.	0.	0.	1.	1.	0.
121	0.	1.	0.	1.	1.	1.	0.	0.	1.	2.
131	2.	0.	1.	3.	1.	0.	0.	1.	0.	1.
141	0.	0.	0.	0.	0.	2.	0.	0.	0.	1.
151	1.	1.	3.	2.	2.	1.	0.	2.	0.	1.
161	1.	0.	2.	0.	1.	1.	1.	4.	1.	2.
171	2.	2.	3.	0.	2.	2.	0.	0.	4.	1.
181	2.	1.	0.	2.	1.	0.	5.	1.	0.	0.
191	0.	0.	3.	1.	1.	1.	2.	4.	3.	2.
201	7.	3.	7.	2.	4.	3.	1.	4.	5.	8.
211	8.	4.	1.	7.	7.	5.	5.	6.	7.	5.
221	9.	3.	6.	11.	8.	12.	6.	16.	20.	27.
231	27.	28.	37.	18.	16.	10.	7.	5.	1.	2.
241	2.	10.	3.	3.	11.	2.	7.	5.	3.	7.
251	5.	6.	5.	2.	7.	4.	5.	5.	8.	6.
261	5.	5.	10.	13.	6.	10.	12.	12.	6.	13.
271	19.	28.	25.	23.	23.	17.	19.	18.	19.	16.
281	10.	17.	23.	27.	28.	34.	46.	42.	51.	73.
291	54.	36.	29.	16.	15.	17.	11.	9.	16.	18.
301	15.	15.	19.	18.	18.	15.	26.	23.	22.	22.
311	24.	10.	17.	17.	15.	16.	21.	17.	22.	29.
321	26.	41.	27.	27.	34.	24.	35.	44.	55.	52.
331	52.	66.	60.	57.	76.	71.	67.	61.	90.	87.
341	87.	92.	106.	139.	139.	145.	159.	202.	199.	209.
351	245.	291.	379.	469.	509.	494.	445.	398.	415.	468.
361	531.	687.	842.	992.	1037.	866.	452.	156.	26.	6.
371	1.	2.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
401	0.	0.	1.	0.	1.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	2.	0.	0.	0.
421	0.	0.	1.	1.	0.	0.	1.	0.	0.	1.
431	3.	1.	0.	3.	1.	1.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	1.	0.	0.	1.
451	0.	0.	1.	0.	0.	1.	1.	2.	0.	0.
461	1.	0.	1.	3.	0.	2.	0.	3.	0.	0.
471	1.	1.	0.	5.	2.	3.	5.	6.	5.	3.
481	5.	2.	0.	0.	1.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S998T002817-DUP-
File ID: 12a1218.CNF

Counted on: 12/ 3/98 @21:29
Detector: AEA12
Geometry number: 1
Count time: 28806. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1102.0	1102.0	360.149	360.149	10.000	2.554	5.000	0.906
2	28.9	28.9	303.095	302.769	58.000	425.974	29.000	124.293
3	35.2	35.2	285.097	285.017	8.000	1.371	4.000	1.289
4?	8.0	8.0	268.757	268.114	6.000	0.898	3.000	6.380
5	23.3	23.3	226.879	226.756	10.000	2.970	5.000	1.247

12544 not available peak JFR

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid Exp.	Obs.	Diff.	FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
1	Pu236	0.846	5.755	5.757	-0.002	0.01	33.24	1.6	151.1	0.681E-04
2	Am241	0.151	5.479	5.493	-0.014	0.96	5.92	9.5	28.1	0.126E-04
3	Pu238		5.487	5.493	-0.006				36.7	0.165E-04
3	Th228	0.014	5.400	5.411	-0.011	0.01	0.53	28.6	3.3	0.151E-05
4	????			5.333			0.08	145.		
5	Pu239	0.016	5.147	5.143	0.004	0.01	0.65	13.3	2.9	0.130E-05
	Pu240		5.144	5.143	0.001				2.9	0.130E-05

Totals: 1.026 <--valid peaks only--> 40.34

DETECTOR CALIBRATION

Energy(MEV) = 4.100 + (0.0046)*Channel
Energy range (MeV): 4.100 TO 6.455
Efficiency = 0.2244 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18873.0	100.000
Smoothed	18873.0	100.000
Composite fit	19403.7	102.812
Residuals	-530.7	-2.812

4863

Analyzed by: _____

Spectrum 12a1218.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 6025.5

5
 5
 5.
 2
 2
 2
 2
 4.
 32
 .3..
 .2
 .2
 .2
 12.
 .21..
 .2....1.
 2.....1
 2.....1....
 2.....
 2.....1.
 2.
 2

Raw Data Dump for AEA Spectrum: 12a1218.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	3.	1.
21	0.	0.	0.	1.	0.	1.	0.	0.	0.	0.
31	0.	0.	1.	0.	0.	2.	0.	0.	1.	1.
41	0.	0.	0.	1.	0.	1.	2.	0.	1.	0.
51	1.	0.	1.	0.	0.	1.	0.	1.	1.	1.
61	2.	0.	0.	2.	0.	0.	0.	0.	0.	0.
71	0.	1.	0.	3.	0.	1.	0.	0.	1.	0.
81	1.	2.	0.	2.	0.	0.	1.	1.	0.	2.
91	0.	1.	0.	5.	0.	0.	1.	1.	0.	1.
101	0.	0.	0.	0.	0.	0.	0.	1.	1.	1.
111	1.	2.	0.	0.	1.	1.	1.	1.	0.	0.
121	1.	0.	0.	1.	1.	1.	0.	0.	3.	0.
131	2.	2.	1.	0.	1.	1.	1.	0.	3.	2.
141	1.	1.	1.	3.	2.	1.	1.	1.	0.	0.
151	0.	0.	2.	0.	1.	5.	2.	0.	2.	2.
161	2.	2.	0.	1.	2.	4.	2.	3.	1.	4.
171	3.	2.	3.	0.	3.	2.	3.	0.	0.	3.
181	0.	1.	2.	1.	6.	1.	4.	4.	5.	2.
191	5.	1.	3.	2.	4.	5.	3.	1.	3.	4.
201	2.	7.	7.	4.	10.	7.	5.	4.	3.	9.
211	5.	6.	7.	10.	4.	10.	12.	17.	14.	14.
221	19.	16.	20.	22.	24.	30.	33.	34.	23.	14.
231	11.	12.	5.	4.	6.	6.	7.	8.	8.	2.
241	5.	11.	6.	3.	7.	6.	8.	6.	10.	7.
251	5.	16.	12.	9.	3.	8.	13.	15.	9.	14.
261	16.	15.	18.	22.	15.	35.	24.	25.	39.	29.
271	18.	25.	19.	17.	24.	16.	26.	44.	38.	31.
281	44.	45.	61.	65.	75.	65.	74.	27.	26.	13.
291	18.	11.	18.	22.	25.	25.	23.	20.	21.	24.
301	37.	37.	36.	37.	34.	23.	28.	28.	23.	35.
311	36.	32.	36.	38.	37.	46.	46.	43.	51.	45.
321	51.	61.	58.	59.	76.	70.	78.	66.	77.	89.
331	78.	95.	77.	115.	110.	112.	122.	145.	136.	141.
341	170.	192.	227.	240.	249.	334.	336.	414.	541.	604.
351	677.	647.	559.	536.	576.	594.	684.	906.	1121.	1353.
361	1407.	1091.	536.	192.	24.	1.	0.	0.	0.	0.
371	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	1.	0.	1.	0.	0.	0.	1.	1.	0.
421	0.	0.	0.	1.	1.	0.	1.	0.	0.	2.
431	1.	1.	0.	0.	0.	0.	0.	0.	1.	0.
441	0.	0.	1.	0.	1.	0.	1.	1.	0.	0.
451	0.	2.	0.	0.	0.	0.	0.	1.	1.	0.
461	0.	0.	3.	0.	1.	3.	1.	2.	2.	5.
471	5.	4.	7.	7.	4.	3.	6.	3.	2.	3.
481	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

LABCORE Completed Worklist Report for Worklist# 26722

Analyst: crj Instrument: AB16 Book#: _____

Method: LA-953-104 Rev/Mod _____

Worklist Comment: BY-105 FOR @PU23901 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@PU23901 PU23901 LIQUID		1.12E-04	1.18E-4	105.357 % Recovery	
1 STD	0	@PU23901 PU23901T LIQUID		100	8.31E+01	83.100 % Recovery	
1 STD	0	@PU23901 PU23901R LIQUID		1.00	2.03E+00	2.030 % Ct Error	
2 BLNK	0	@PU23901 PU23901 LIQUID		1	<3.35E-4		uCi/mL
2 BLNK	0	@PU23901 PU23901T LIQUID		100	8.60E+01	86.000 % Recovery	
2 BLNK	0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000	uCi/mL
3 SAMPLE	S98T002818 0	@PU23901 PU23901 LIQUID		N/A	<3.49E-4	3.49E-004	uCi/mL
3 SAMPLE	S98T002818 0	@PU23901 PU23901T LIQUID		N/A	8.51E+01		% Recovery
3 SAMPLE	S98T002818 0	@PU23901 PU23901R LIQUID		N/A	1.04E+01		% Ct Error
4 DUP	S98T002818 0	@PU23901 PU23901 LIQUID		<3.49E-4	<3.47E-4		RPD
4 DUP	S98T002818 0	@PU23901 PU23901T LIQUID		100	8.44E+01	84.400 % Recovery	
4 DUP	S98T002818 0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000 % Ct Error	
5 SAMPLE	S98T002819 0	@PU23901 PU23901 LIQUID		N/A	<3.47E-4	3.47E-004	uCi/mL
5 SAMPLE	S98T002819 0	@PU23901 PU23901T LIQUID		N/A	8.34E+01		% Recovery
5 SAMPLE	S98T002819 0	@PU23901 PU23901R LIQUID		N/A	1.00E+02		% Ct Error
6 DUP	S98T002819 0	@PU23901 PU23901 LIQUID		<3.47E-4	<3.47E-4		RPD
6 DUP	S98T002819 0	@PU23901 PU23901T LIQUID		100	8.29E+01	82.900 % Recovery	
6 DUP	S98T002819 0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000 % Ct Error	
7 SAMPLE	S98T002820 0	@PU23901 PU23901 LIQUID		N/A	<3.43E-4	3.43E-004	uCi/mL
7 SAMPLE	S98T002820 0	@PU23901 PU23901T LIQUID		N/A	8.38E+01		% Recovery
7 SAMPLE	S98T002820 0	@PU23901 PU23901R LIQUID		N/A	8.38E+00		% Ct Error
8 DUP	S98T002820 0	@PU23901 PU23901 LIQUID		<3.43E-4	<3.69E-4		RPD
8 DUP	S98T002820 0	@PU23901 PU23901T LIQUID		100	8.09E+01	80.900 % Recovery	
8 DUP	S98T002820 0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000 % Ct Error	
9 SAMPLE	S98T002821 0	@PU23901 PU23901 LIQUID		N/A	<3.50E-4	3.50E-004	uCi/mL
9 SAMPLE	S98T002821 0	@PU23901 PU23901T LIQUID		N/A	8.43E+01		% Recovery
9 SAMPLE	S98T002821 0	@PU23901 PU23901R LIQUID		N/A	1.00E+02		% Ct Error
10 DUP	S98T002821 0	@PU23901 PU23901 LIQUID		<3.50E-4	<3.40E-4		RPD
10 DUP	S98T002821 0	@PU23901 PU23901T LIQUID		100	8.55E+01	85.500 % Recovery	
10 DUP	S98T002821 0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000 % Ct Error	
11 SAMPLE	S98T002939 0	@PU23901 PU23901 LIQUID		N/A	<3.12E-4	3.12E-004	uCi/mL
11 SAMPLE	S98T002939 0	@PU23901 PU23901T LIQUID		N/A	8.76E+01		% Recovery
11 SAMPLE	S98T002939 0	@PU23901 PU23901R LIQUID		N/A	9.25E+00		% Ct Error
12 DUP	S98T002939 0	@PU23901 PU23901 LIQUID		<3.12E-4	<2.94E-4		RPD
12 DUP	S98T002939 0	@PU23901 PU23901T LIQUID		100	8.99E+01	89.900 % Recovery	
12 DUP	S98T002939 0	@PU23901 PU23901R LIQUID		1.00	1.00E+02	100.000 % Ct Error	

Final page for worklist# 26722

LABCORE Completed Worklist Report for Worklist# 26722

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
Analyst Signature		Date					
Analyst Signature		Date					

<u>John Relyea</u>	<u>7 Dec 98</u>
Reviewer Signature	Date

LABCORE Data Entry Template for Worklist# 26722

Analyst: CRT Instrument: PU01 14 Book# 28BS7

Method: LA-953-104 Rev/Mod B-D

Worklist Comment: BY-105 FOR @PU23901 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@PU23901	LIQUID		
2 BLNK			@PU23901	LIQUID		
3 SAMPLE	S98T002818 0		@PU23901	LIQUID	98000437	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
4 DUP	S98T002818 0		@PU23901	LIQUID		
5 SAMPLE	S98T002819 0		@PU23901	LIQUID	98000437	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
6 DUP	S98T002819 0		@PU23901	LIQUID		
7 SAMPLE	S98T002820 0		@PU23901	LIQUID	98000437	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
8 DUP	S98T002820 0		@PU23901	LIQUID		
9 SAMPLE	S98T002821 0		@PU23901	LIQUID	98000437	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
10 DUP	S98T002821 0		@PU23901	LIQUID		
11 SAMPLE	S98T002939 0		@PU23901	LIQUID	98000490	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
12 DUP	S98T002939 0		@PU23901	LIQUID		

Final page for worklist # 26722

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID****STD**

Type	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	0.353	
STD	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.034
Work List	SAMPLE DILUTION FACTOR	DF	1.000	PU 239 AEA FRAC (C239)	0.504
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	3524
Test Code	DIGEST DILUTION FACTOR	DDF	1.000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130	
Matrix	DETECTOR NUMBER	16	PU 236 cpm	33.100	
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm	3.150
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	47.250	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480	
ReRun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L	1.1805E-01	
0	PU-238 TRACER VALUE (dpm/mL)	0.000			
Sample Prep	STANDARD BOOK NO	28B57			
N/A	STANDARD VALUE in μ Ci/mL	1.120E-04			
Sample #					
WL26722					
Instrument Code	Decay Time = Date Counted - Tracer Preparation Date				
WB27806	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * e to the power of $[-(\ln 2 * \text{Decay Time}/1040.95)]$				
Prepared By	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236*100/Pu-236 Decay Corr'd Value * SPKV				
SZC	Pu 239/240 μ Ci/L = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)/(1000mL/L)(DF)(DDF) / [(C236)(SS)(2220000 dpm/ μ Ci)]				
Chemist	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)				
JFR	Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]				
Analyst	Relative Counting Error = Square Root of $[(1/(\text{Pu 236 cpm} * \text{min})) + (1/(\text{Pu 238 or 239/240 cpm} * \text{min}))] * 1.96 * 100$				
CRJ					
Date Complete					
12/07/98	Pu 239/240 μ Ci/mL	1.18E-04	DETECTION LEVELS in μ Ci/mL Pu 239/240 8.46E-06		
Analysis Date	Relative Counting Error =	2.0%			
12/04/98					
Analysis Time					
02:45 PM					
Sample Point					
BY-105	Pu 236 Tracer Recovery =	83.1%			

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
STANDARD.WB1 REV 1.0	943128ML		

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LI

BLNK

Type	DATE COUNTED	DEC-04-98	Pu 236 AEA FRAC (C236)	0.871
BLNK	SAMPLE VOLUME in mL	SS 1.000	Pu 238 AEA FRAC (C238)	0.000
Worklist	SAMPLE DILUTION FACTOR	DF 101.000	Pu 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL	SPKV 0.100	TOTAL AT COUNTS	1481
Test Code	DIGEST DILUTION FACTOR	DDF 1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cp (Bkg)	0.130
Matrix	DETECTOR NUMBER	16	Pu 236 cpm	33.770
LIQUID	EFFICIENCY FACTOR	EFF 0.2717	Pu 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	0.000
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.00	AEA COUNT TIME	480
Rorun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.59	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.345E-01
0	PU-238 TRACER VALUE (dpm/m)	0.00		

Sample Prep
N/ASample #
WL26722

Decay Time = Date Counted - Tracer Preparation Date

Instrument Code
WB27806

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Prepared By
SZC

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Chemist
JFRPu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μCi)]Analyst
CRJ

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Date Complete
12/07/98Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) * (DF) * (DDF) / (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μCi) * (D g/L) * (SS)]Analysis Date
12/04/98

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min))] + (1 / (Pu 238 or 239/240 cpm * min)) * 1.96 * 100

Analysis Time
02:45 PMSample Point
BY-105

Pu 239/240 $\mu\text{Ci/mL}$	< 3.35E-04	DETECTION LEVELS in $\mu\text{Ci/mL}$
Relative Counting Error	= 100.0%	
NOTE: Pu 238 Result is a LESS THAN Value.		
Pu 238 $\mu\text{Ci/mL}$	< 3.35E-04	Pu 239/240
Relative Counting Error	= 100.0%	3.35E-04
Pu 236 Tracer Recovery	= 86.0%	Pu 238
		3.35E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

BLANK.WB1 REV 1.0

943126ML

4870

12/07/98

WORKBOOK PAGE: SAM3

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	SAMPLE
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)
WORK LIST	SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC (C239)
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS
Tracer Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	
Matrix	DETECTOR NUMBER	16	PU 236 cpm	
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	
Re-run	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L =	< 3.4856E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002818

Instrument Code

WB27806

Prepared By

SZC

Chemist

JFR

Analyst

CRJ

Date Completed

12/07/98

Analysis Date

12/04/98

Analysis Time

02:45 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 μ Ci/mL	< 3.49E-04	DETECTION LEVELS in μ Ci/mL
Analysis Date	Relative Counting Error	= 10.4%	
12/04/98			Pu 239/240
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		3.49E-04
02:45 PM	Pu 238 μ Ci/mL	< 3.49E-04	Pu 238
Sample Point	Relative Counting Error	= 100.0%	3.49E-04
BY-105	Pu 236 Tracer Recovery	= 85.1%	

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4871

WORKBOOK PAGE: DUP4

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	0.856	
DUP	SAMPLE VOLUME in mL SS	1.000	PU 238 AEA FRAC (C238)	0.057
SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	1479
DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130
DETECTOR NUMBER	16	PU 236 cpm	32.130	
LIQUID	EFFICIENCY FACTOR EFF	0.272	PU 238 cpm	2.130
TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
PU-238 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L =	<	3.4680E-01
0	PU-238 TRACER VALUE (dpm/mL)			

Sample Pre

N/A

Sample #

S98T002818

Instrument Code

WB27806

Prepared By

SZC

Chemist

JFR

Analyst

CRJ

Date Complete

12/07/98

Analysis Date

12/04/98

Analysis Time

02:45 PM

Sample Bottle

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(t * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) / (1000 mL)] / [(Pu-238 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 μ Ci/mL	< 3.47E-04	DETECTION LEVELS
Analysis Date	Relative Counting Error	= 100.0%	in μ Ci/mL
12/04/98			Pu 239/240
Analysis Time	Pu 238 μ Ci/mL	5.56E-04	3.47E-04
02:45 PM	Relative Counting Error	= 6.3%	Pu 238
Sample Bottle	Pu 236 Tracer Recovery	= 84.4%	3.47E-04
BY-105			

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
SAMPLE WB1 REV 1.0	943128ML		

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /**

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))				LIQUID /	SAMPLE	
Type	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC	(C236)	0.865	
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC	(C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC	(C239)	0.000
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS		1447
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)		30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.130	
Matrix	DETECTOR NUMBER	16	PU 236 cpm		33.450	
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm		0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.000	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480	
Re-run	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L =		< 3.4715E-01	
0	PU-238 TRACER VALUE (dpm/mL)	0.000				

Sample Prep

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

S98T002819

Instrument Code

WB27806

Prepared By

SZC

Chemist

JFR

Analyst

CRJ

Date Complete

12/07/98

Analysis Date

12/04/98

Analysis Time

02:45 PM

Sample Point

BY-105

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 μ Ci/mL	< 3.47E-04	DETECTION LEVELS
Analysis Date	Relative Counting Error	= 100.0%	in μ Ci/mL
12/04/98			Pu 239/240
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		3.47E-04
02:45 PM	Pu 238 μ Ci/mL	< 3.47E-04	Pu 238
Sample Point	Relative Counting Error	= 100.0%	3.47E-04
BY-105	Pu 236 Tracer Recovery	= 83.4%	

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98
SAMPLE WB1 REV 1.0	943128ML		

4873

WORKBOOK PAGE: DUP6

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

Activity	DATE COUNTED	DEC-04-98	Pu 236 AEA FRAC (C236)	0.871
DUP	SAMPLE VOLUME in mL	SS	1.000 Pu 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	101.000 Pu 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1428
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130
Main	DETECTOR NUMBER	16	Pu 236 cpm	32.780
LIQUID	EFFICIENCY FACTOR	EFF	0.272 Pu 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	0.000
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Rerun	Pu-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 µCi/L =	< 3.4696E-01
0	Pu-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002819

Instrument Code

WB27806

Prepared By

SZC

Chemist

JFR

Analyst

CRJ

Date Complete

12/07/98

Analysis Date

12/04/98

Analysis Time

02:45 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 µCi/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/µCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 µCi/L = [(Pu 238 dpm) * (DF) * (DDF) / (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/µCi) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 µCi/mL	< 3.47E-04	DETECTION LEVELS in µCi/mL
	Relative Counting Error	= 100.0%	
12/04/98	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
02:45 PM	Pu 238 µCi/mL	< 3.47E-04	3.47E-04
	Relative Counting Error	= 100.0%	Pu 238
	Pu 236 Tracer Recovery	= 82.9%	3.47E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	<i>John Relyea</i>	Date:	7 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

WORKBOOK PAGE: SAM7

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

Type	DATE COUNTED	DEC-04-98	PU 238 AEA FRAC (C238)	0.871
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)
WORKLIST	SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC (C239)
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130
Magn	DETECTOR NUMBER	16	PU 236 cpm	33.770
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.180
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Recovery	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 µCi/L =	< 3.4334E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep**N/A****Sample #****S98T002820****Instrument Code****WB27806****Prepared By****SZC****Chemist****JFR****Analyst****CRJ****Date Complete****12/07/98****Analysis Date****12/04/98****Analysis Time****02:45 PM****Sample Point****BY-105**

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 µCi/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV / (1000mL/L)) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/µCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 µCi/L = [(Pu 238 dpm) * (DDF) * (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) * (2220000 dpm/µCi) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 µCi/mL	< 3.43E-04	DETECTION LEVELS
Analysis Date	Relative Counting Error	= 8.4%	in µCi/mL
12/04/98			Pu 239/240
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		3.43E-04
02:45 PM	Pu 238 µCi/mL	< 3.43E-04	Pu 238
Sample Point	Relative Counting Error	= 100.0%	3.43E-04
BY-105	Pu 236 Tracer Recovery	= 83.8%	

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE WB1 REV 1.0

943128ML

4875

WORKBOOK PAGE: DUP8

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

Type	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	0.839	
DUP	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.000
Work Dist	SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1447
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130	
Main	DETECTOR NUMBER	16	PU 236 cpm	31.060	
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480	
Rerun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L =	<	3.6900E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample #

S98T002820

Decay Time = Date Counted - Tracer Preparation Date

Instrument Code

WB27806

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln 2 * Decay Time)/1040.95}]

Prepared By

SZC

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Chemist

JFR

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000mL/L) * (DF) * (DDF) / [(C238) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Analyst

CRJ

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

Date Complete

12/07/98

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Analysis Date

12/04/98

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Analysis Time

02:45 PM

NOTE: Pu 238 Result is a LESS THAN Value.

Sample Point

BY-105

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Date

12/04/98

DETECTION LEVELS in μ Ci/mL

Analysis Time

02:45 PM

Pu 239/240 μ Ci/mL

Sample Point

BY-105

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

Sample Point

BY-105

DETECTION LEVELS in μ Ci/mL

Date

12/07/98

Pu 239/240 μ Ci/mL

Analysis Time

02:45 PM

Pu 238 μ Ci/mL

Sample Point

BY-105

Relative Counting Error = 100.0%

Date

12/07/98

Pu 236 Tracer Recovery = 80.9%

Analysis Time

02:45 PM

DETECTION LEVELS in μ Ci/mL

Sample Point

BY-105

Pu 239/240 μ Ci/mL

Date

12/07/98

Pu 238 μ Ci/mL

Analysis Time

02:45 PM

Relative Counting Error = 100.0%

Sample Point

BY-105

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

DETECTION LEVELS in μ Ci/mL

Analysis Time

02:45 PM

Pu 239/240 μ Ci/mL

Sample Point

BY-105

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

DETECTION LEVELS in μ Ci/mL

Sample Point

BY-105

Pu 239/240 μ Ci/mL

Analysis Time

02:45 PM

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

DETECTION LEVELS in μ Ci/mL

Sample Point

BY-105

Pu 239/240 μ Ci/mL

Analysis Time

02:45 PM

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

DETECTION LEVELS in μ Ci/mL

Sample Point

BY-105

Pu 239/240 μ Ci/mL

Analysis Time

02:45 PM

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

Date

12/07/98

DETECTION LEVELS in μ Ci/mL

Sample Point

BY-105

Pu 239/240 μ Ci/mL

Analysis Time

02:45 PM

Pu 238 μ Ci/mL

Date

12/07/98

Relative Counting Error = 100.0%

Analysis Time

02:45 PM

Pu 236 Tracer Recovery = 80.9%

4876

12/07/98

WORKBOOK PAGE: SAM9

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

SAMPLE

TYPE	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	0.849	
SAMPLE	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.000
WORKLIST	SAMPLE DILUTION FACTOR	DF	101.000	PU 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1489
TEST CODE	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130	
MATRIX	DETECTOR NUMBER	16	PU 236 cpm	32.250	
LIQUID	EFFICIENCY FACTOR	EFF	0.272	PU 238 cpm	0.000
BATCH NUMBER	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000	
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480	
RETURN	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 $\mu\text{Ci/L}$ =	<	3.5017E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Preparation

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

S98T002821

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WB27806

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μCi)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

SZC

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μCi) * (D g/L) * (SS)]

Chemist

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

CRJ

Data Complete

12/07/98

Pu 239/240 $\mu\text{Ci/mL}$ < 3.50E-04

Analysis Date

Relative Counting Error = 100.0%

12/04/98

Analysis Time

NOTE: Pu 238 Result is a LESS THAN Value.

02:45 PM

Pu 238 $\mu\text{Ci/mL}$ < 3.50E-04

Sample Point

Relative Counting Error = 100.0%

BY-105

Pu 236 Tracer Recovery = 84.3%

DETECTION

LEVELS

in $\mu\text{Ci/mL}$

Pu 239/240

3.50E-04

Pu 238

3.50E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: DUP10

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

DATE COUNTED	DEC-04-98	PU 238 AEA FRAC (C238)	0.881
DUP	SAMPLE VOLUME in mL	SS	1.000
WORKLIST	SAMPLE DILUTION FACTOR	DF	101.000
26722	TRACER VOLUME in mL	SPKV	0.100
TEST CODE	DIGEST DILUTION FACTOR	DDF	1.0000
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)
MAINT	DETECTOR NUMBER	16	PU 236 cpm
LIQUID	EFFICIENCY FACTOR	EFF	0.272
BATCH NUMBER	TRACER PREPARATION DATE	01/19/98	PU 239 cpm
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME
REMARKS	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 μ Ci/L =
0	PU-238 TRACER VALUE (dpm/mL)	0.000	< 3.4025E-01
SAMPLE PREP			
N/A			

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV / (1000mL/L)) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) / (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

ANALYST

CRJ

DATE COMPLETE

12/07/98

ANALYSIS DATE

12/04/98

ANALYSIS TIME

02:45 PM

SAMPLE POINT

BY-105

Pu 239/240 μ Ci/mL

Relative Counting Error =

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 238 μ Ci/mL

Relative Counting Error =

Pu 236 Tracer Recovery =

DETECTION

LEVELS

in μ Ci/mL

Pu 239/240

3.40E-04

Pu 238

3.40E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE WB1 REV 1.0

943128ML

4878

WORKBOOK PAGE: SAM11

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

TYPE	DATE COUNTED	DEC-04-98	PU 236 AEA FRAC (C236)	0.917
SAMPLE	SAMPLE VOLUME in mL	SS	1.000 PU 238 AEA FRAC (C238)	0.000
WORKLIST	SAMPLE DILUTION FACTOR	DF	101.000 PU 239 AEA FRAC (C239)	0.025
26722	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1433
TEST CODE	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130
MATRIX	DETECTOR NUMBER	16	PU 236 cpm	35.090
LIQUID	EFFICIENCY FACTOR	EFF	0.272 PU 238 cpm	0.000
BATCH NUMBER	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.960
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
REFURIN	PU-236 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.1192E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

SAMPLE PREP

N/A

Decay Time = Date Counted - Tracer Preparation Date

S98T002939 Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln 2 * Decay Time/1040.95)}]

INSTRUMENT CODE Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WB27806 Pu 239/240 $\mu\text{Ci/L}$ = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Prepared By Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

SZC Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

CHEMIST Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

ANALYST

CRJ

DATE COMPLETE

12/07/98

ANALYSIS DATE

12/04/98

ANALYSIS TIME

02:45 PM

SAMPLE POINT

BY-105

Pu 239/240 $\mu\text{Ci/mL}$	< 3.12E-04	DETECTION LEVELS in $\mu\text{Ci/mL}$
Relative Counting Error =	9.3%	
NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
Pu 238 $\mu\text{Ci/mL}$	< 3.12E-04	3.12E-04
Relative Counting Error =	100.0%	Pu 238
Pu 236 Tracer Recovery =	87.6%	3.12E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4879

WORKBOOK PAGE: DUP12

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

Type	DATE COUNTED	DEC-04-98	PU 238 AEA FRAC (C236)	0.950
DUP	SAMPLE VOLUME in mL	SS 1.000	PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF 101.000	PU 239 AEA FRAC (C239)	0.000
26722	TRACER VOLUME in mL	SPKV 0.100	TOTAL AT COUNTS	1419
Test Code	DIGEST DILUTION FACTOR	DDF 1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.130
Matrix	DETECTOR NUMBER	16	PU 236 cpm	34.670
LIQUID	EFFICIENCY FACTOR	EFF 0.272	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
98005328	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Return	PU-238 DECAY CORR'D VALUE (dpm/mL)	1835.587	Pu 239/240 $\mu\text{Ci/L}$ =	< 2.9351E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002939

Instrument Code

WB27806

Prepared By

SZC

Chemist

JFR

Analyst

CRJ

Date Complete

12/07/98

Analysis Date

12/04/98

Analysis Time

02:45 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) (SPKV) / (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) / (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/07/98	Pu 239/240 $\mu\text{Ci/mL}$	< 2.94E-04	DETECTION LEVELS in $\mu\text{Ci/mL}$
Analysis Date	Relative Counting Error	= 100.0%	
12/04/98	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
Analysis Time	Pu 238 $\mu\text{Ci/mL}$	< 2.94E-04	2.94E-04
02:45 PM	Relative Counting Error	= 100.0%	Pu 238
Sample Point	Pu 236 Tracer Recovery	= 89.9%	2.94E-04

Analyst:	CRJ	Date:	07-Dec-98
Signature of Chemist:	JFR	Date:	7 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4880

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL26722-STD-PU
 File ID: 1a1866.CNF

Counted on: 12/ 4/98 @22:45
 Detector: AEA1
 Geometry number: 1
 Count time: 28805. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	797.6	797.6	361.734	361.734	10.000	2.983	5.000	0.699
2	31.1	31.1	303.472	302.666	12.000	12.965	6.000	1.190
3	36.0	36.0	287.795	287.475	8.000	2.293	4.000	0.865
4?	14.2	14.2	270.734	267.830	12.000	4.618	6.000	2.810
5	1339.7	1339.7	229.361	229.360	10.000	3.125	5.000	0.924
6	13.0	13.0	120.731	120.704	200.000	1.000	100.000	0.100
7?	2.5	2.5	111.895	111.122	4.000	0.438	2.000	0.551

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid Exp.	Obs.	Diff.	FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
1	Pu236	0.353	5.755	5.754	0.0010	0.01	33.10	1.6	140.6	0.633E-04
2	Am241	0.034	5.479	5.482	-0.0030	0.06	3.15	6.7	13.9	0.628E-05
	Pu238		5.487	5.482	0.005				18.2	0.820E-05
3	Th228	0.011	5.400	5.412	-0.0120	0.01	1.02	16.8	6.0	0.270E-05
4		????		5.322			0.37	38.1		
5	Pu239	0.504	5.147	5.145	0.0020	0.01	47.25	1.3	196.7	0.886E-04
	Pu240		5.144	5.145	-0.001				196.7	0.886E-04
6	Np237	0.011	4.640	4.645	-0.0050	0.00	1.01	9.1	70.1	0.316E-04
7		????		4.601			0.02	340.		
Totals:		0.913	<--valid peaks only-->				85.53			

DETECTOR CALIBRATION

Energy(MEV) = 4.090 + (0.0046)*Channel
 Energy range (MeV): 4.090 TO 6.445
 Efficiency = 0.2402 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	44983.0	100.000
Smoothed	44982.6	99.999
Composite fit	44881.41248.7	91.698

Residuals

3734.3

8.302

Analyzed by:

SLH2

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 9070.6

```

6
6
6
6
6
6
6
7
6
6
.
.
..
..
...
5...
..5....
....5....
.....5...
.....5..
2.....5
2.....5
2.....5..
2
2
4
4.
4.
3..
3.
2.
2.
21.
...1.
.....1
.....1
.....1.
..... 1
.....1
..

```

Raw Data Dump for AEA Spectrum: 1a1866.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	2.	2.	3.
21	2.	3.	2.	0.	2.	3.	1.	6.	2.	3.
31	3.	0.	5.	2.	3.	2.	2.	2.	7.	3.
41	1.	5.	0.	3.	4.	2.	5.	6.	7.	6.
51	3.	5.	5.	3.	6.	1.	5.	0.	4.	6.
61	5.	4.	7.	3.	5.	10.	7.	5.	5.	6.
71	4.	2.	4.	10.	5.	5.	4.	3.	6.	5.
81	6.	10.	4.	6.	6.	6.	3.	4.	9.	6.
91	9.	4.	7.	5.	12.	8.	8.	2.	9.	8.
101	6.	11.	9.	8.	8.	12.	10.	10.	13.	10.
111	11.	18.	12.	10.	9.	13.	9.	6.	12.	17.
121	15.	16.	8.	11.	12.	11.	9.	10.	10.	13.
131	19.	15.	16.	20.	21.	19.	18.	12.	24.	24.
141	23.	27.	28.	25.	23.	21.	18.	28.	32.	27.
151	26.	28.	30.	35.	27.	32.	25.	48.	37.	38.
161	30.	45.	41.	38.	41.	52.	48.	52.	51.	43.
171	53.	53.	55.	68.	74.	74.	51.	73.	67.	90.
181	93.	86.	80.	113.	93.	108.	113.	113.	107.	131.
191	138.	140.	130.	169.	161.	169.	172.	191.	193.	207.
201	209.	198.	192.	248.	252.	271.	289.	322.	312.	341.
211	373.	396.	416.	459.	466.	536.	569.	595.	640.	664.
221	721.	757.	786.	862.	909.	1048.	1147.	1311.	1516.	1627.
231	1479.	965.	500.	151.	34.	12.	9.	13.	6.	16.
241	13.	13.	12.	15.	8.	11.	10.	10.	11.	11.
251	17.	10.	17.	14.	13.	13.	20.	16.	21.	18.
261	13.	22.	18.	22.	27.	40.	38.	36.	39.	35.
271	36.	38.	31.	42.	23.	26.	20.	50.	36.	46.
281	43.	39.	58.	46.	58.	55.	68.	76.	74.	51.
291	44.	35.	32.	34.	33.	46.	41.	32.	36.	46.
301	42.	48.	55.	57.	46.	41.	45.	32.	37.	46.
311	37.	50.	37.	44.	59.	63.	72.	60.	65.	61.
321	77.	83.	78.	92.	105.	98.	110.	119.	119.	117.
331	130.	123.	138.	150.	133.	151.	149.	175.	245.	187.
341	223.	237.	220.	264.	283.	289.	319.	343.	413.	406.
351	484.	607.	597.	545.	508.	508.	481.	558.	623.	752.
361	784.	947.	953.	697.	348.	125.	27.	5.	1.	1.
371	1.	2.	0.	1.	0.	0.	0.	0.	0.	1.
381	1.	0.	2.	3.	1.	0.	0.	0.	1.	0.
391	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
401	0.	2.	0.	0.	1.	0.	1.	2.	0.	1.
411	1.	1.	0.	0.	1.	0.	0.	2.	2.	0.
421	0.	0.	1.	1.	0.	0.	0.	2.	1.	1.
431	1.	1.	0.	1.	1.	1.	0.	0.	0.	1.
441	0.	0.	1.	2.	1.	0.	0.	1.	0.	2.
451	0.	1.	0.	1.	2.	1.	2.	3.	0.	1.
461	0.	0.	1.	2.	0.	1.	1.	2.	2.	2.
471	5.	3.	5.	4.	3.	5.	3.	4.	1.	0.
481	4.	1.	0.	0.	0.	0.	0.	0.	0.	1.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL26722-BLNK-PU
 File ID: 2a2064.CNF

Counted on: 12/ 4/98 @22:45
 Detector: AEA2
 Geometry number: 1
 Count time: 28810. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	839.6	839.6	362.256	362.256	10.000	3.195	5.000	0.790
2	62.2	62.2	288.136	288.039	16.000	13.218	8.000	2.757

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate c/m	%err @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.				
1	Pu236	0.871	5.755	5.754	0.0010.01	33.77	1.5	147.4	0.664E-04
2	Th228	0.092	5.400	5.413	-0.0130.06	3.55	5.2	21.4	0.963E-05
Totals:		0.962	<--valid peaks only-->			37.32			

DETECTOR CALIBRATION

Energy(MEV) = 4.088 + (0.0046)*Channel
 Energy range (MeV): 4.088 TO 6.443
 Efficiency = 0.2338 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18623.0	100.000
Smoothed	18622.8	99.999
Composite fit	17917.7	96.213
Residuals	705.3	3.787

Analyzed by:

SLH2

Spectrum 2a2064.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc   Display Max.:   4712.2

```

.
2
2.
.2.
...2
.....2
....2
2...
.1..
...1.
.....1.
.....1
..... 1
.....1..
..... 1
.....1
..1...

```

Raw Data Dump for AEA Spectrum: 2a2064.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	1.	0.	0.	0.	0.	0.	0.	0.	1.
31	0.	0.	1.	1.	0.	0.	0.	0.	0.	1.
41	0.	2.	1.	0.	0.	0.	0.	3.	0.	1.
51	0.	1.	2.	0.	0.	0.	0.	1.	0.	0.
61	0.	0.	0.	3.	0.	2.	0.	1.	0.	0.
71	0.	1.	1.	0.	1.	0.	2.	2.	2.	1.
81	3.	1.	0.	0.	2.	1.	0.	2.	2.	2.
91	2.	3.	1.	1.	0.	1.	2.	0.	2.	0.
101	1.	2.	0.	0.	0.	0.	1.	2.	0.	0.
111	0.	1.	3.	1.	1.	0.	0.	2.	1.	0.
121	2.	0.	2.	1.	1.	0.	1.	1.	2.	1.
131	1.	1.	4.	5.	1.	5.	1.	2.	2.	1.
141	1.	1.	2.	1.	1.	1.	1.	0.	4.	0.
151	4.	1.	3.	2.	1.	2.	2.	2.	3.	1.
161	3.	3.	0.	1.	1.	6.	4.	2.	1.	1.
171	1.	4.	1.	3.	2.	4.	3.	1.	3.	1.
181	2.	1.	3.	4.	0.	3.	0.	0.	0.	4.
191	3.	1.	2.	1.	2.	4.	4.	10.	2.	0.
201	4.	5.	0.	2.	4.	2.	3.	3.	8.	2.
211	5.	2.	3.	4.	7.	3.	3.	5.	6.	6.
221	5.	5.	5.	3.	10.	10.	3.	10.	4.	7.
231	7.	15.	5.	7.	4.	12.	4.	9.	8.	4.
241	3.	8.	7.	13.	14.	7.	11.	10.	6.	5.
251	14.	5.	8.	15.	13.	19.	19.	11.	15.	16.
261	18.	15.	27.	20.	25.	26.	27.	18.	34.	31.
271	32.	29.	38.	28.	32.	31.	29.	27.	34.	38.
281	48.	43.	52.	49.	52.	62.	66.	76.	86.	61.
291	46.	30.	38.	35.	36.	32.	20.	22.	35.	29.
301	40.	41.	37.	45.	38.	28.	41.	41.	48.	47.
311	41.	40.	45.	45.	54.	57.	34.	46.	52.	60.
321	66.	58.	89.	74.	76.	66.	81.	98.	112.	104.
331	106.	111.	116.	123.	143.	131.	157.	189.	171.	171.
341	181.	194.	219.	245.	250.	297.	287.	345.	407.	426.
351	489.	533.	584.	562.	537.	518.	557.	612.	642.	733.
361	845.	898.	1019.	909.	593.	270.	84.	18.	3.	1.
371	2.	2.	1.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	1.	0.	0.	0.	1.	1.	0.
391	0.	1.	1.	0.	0.	0.	0.	0.	0.	1.
401	0.	1.	0.	0.	0.	0.	0.	0.	1.	1.
411	0.	0.	1.	0.	0.	2.	2.	0.	0.	0.
421	0.	1.	0.	0.	0.	1.	0.	1.	0.	0.
431	2.	0.	1.	2.	2.	0.	0.	1.	1.	0.
441	0.	0.	0.	1.	1.	0.	1.	0.	0.	2.
451	2.	0.	0.	0.	1.	1.	2.	1.	1.	1.
461	1.	1.	1.	0.	0.	2.	0.	4.	0.	1.
471	2.	1.	2.	0.	10.	8.	4.	4.	6.	2.
481	2.	3.	1.	0.	0.	0.	0.	1.	0.	0.
491	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2818-SAM-PU
File ID: 3a3356.CNF

Counted on: 12/ 4/98 @22:47
Detector: AEA3
Geometry number: 1
Count time: 28800. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	948.2	948.2	361.421	361.421	10.000	3.202	5.000	1.018
2	70.0	70.0	286.723	286.693	12.000	7.000	6.000	1.915
3	10.7	10.7	227.460	227.233	16.000	25.933	8.000	4.819

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m	d/m	Activity uCi/ea	
			Exp.	Obs.	Diff.					
1	Pu236	0.845	5.755	5.753	0.0020	0.01	31.93	1.6	139.1	0.627E-04
2	Th228	0.079	5.400	5.409	-0.0090	0.03	2.99	5.3	18.0	0.810E-05
3	Pu239	0.020	5.147	5.136	0.0110	0.12	0.76	11.8	3.2	0.146E-05
	Pu240		5.144	5.136	0.008				3.2	0.146E-05

Totals:		0.945	<--valid peaks only-->				35.68			

DETECTOR CALIBRATION

Energy (MeV) = 4.091 + (0.0046)*Channel

Energy range (MeV): 4.091 TO 6.446

Efficiency = 0.2342 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18131.0	100.000
Smoothed	18131.4	100.002
Composite fit	17125.9	94.456
Residuals	1005.1	5.544

Analyzed by:

SLH2

Raw Data Dump for AEA Spectrum: 3a3356.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	1.	1.	0.
21	1.	1.	0.	0.	1.	1.	0.	3.	0.	0.
31	0.	1.	0.	0.	1.	0.	1.	0.	1.	0.
41	0.	1.	0.	0.	0.	0.	0.	0.	1.	0.
51	0.	1.	0.	0.	0.	1.	1.	0.	0.	1.
61	0.	0.	1.	0.	0.	0.	0.	0.	0.	1.
71	0.	1.	1.	0.	2.	1.	1.	2.	0.	1.
81	1.	0.	0.	1.	1.	0.	0.	0.	0.	2.
91	0.	0.	1.	0.	1.	0.	1.	0.	0.	0.
101	1.	0.	0.	1.	0.	0.	3.	1.	1.	2.
111	0.	1.	1.	3.	0.	3.	1.	2.	1.	2.
121	3.	0.	0.	1.	2.	2.	1.	1.	1.	1.
131	2.	2.	0.	0.	3.	1.	2.	2.	0.	2.
141	0.	1.	0.	0.	2.	3.	1.	2.	2.	1.
151	1.	2.	2.	2.	1.	2.	1.	2.	1.	1.
161	0.	1.	2.	3.	3.	1.	2.	0.	2.	1.
171	0.	2.	2.	1.	1.	3.	2.	0.	0.	1.
181	5.	3.	1.	0.	2.	1.	4.	2.	4.	5.
191	4.	3.	2.	2.	1.	5.	1.	3.	3.	3.
201	4.	0.	2.	3.	4.	6.	6.	7.	9.	4.
211	7.	8.	7.	6.	8.	6.	7.	6.	6.	6.
221	12.	6.	6.	12.	12.	13.	14.	9.	13.	13.
231	8.	3.	10.	3.	8.	7.	7.	5.	8.	5.
241	8.	6.	9.	8.	13.	11.	12.	4.	10.	12.
251	12.	11.	8.	14.	11.	15.	16.	12.	19.	11.
261	19.	18.	19.	19.	24.	25.	26.	31.	35.	37.
271	24.	38.	33.	15.	19.	27.	37.	21.	22.	34.
281	42.	40.	41.	59.	77.	72.	94.	69.	66.	36.
291	32.	22.	25.	19.	16.	22.	18.	21.	27.	25.
301	31.	33.	39.	38.	20.	32.	32.	32.	28.	26.
311	35.	40.	39.	35.	44.	36.	35.	33.	40.	47.
321	47.	54.	53.	56.	63.	61.	69.	69.	66.	89.
331	85.	82.	90.	104.	112.	125.	108.	130.	143.	158.
341	175.	169.	175.	216.	240.	243.	267.	315.	398.	471.
351	497.	592.	606.	534.	537.	517.	605.	654.	791.	972.
361	1103.	1114.	1070.	732.	341.	129.	37.	11.	3.	1.
371	2.	1.	1.	1.	2.	0.	0.	0.	1.	1.
381	0.	0.	0.	0.	0.	0.	2.	1.	1.	1.
391	0.	0.	0.	0.	0.	2.	1.	1.	0.	0.
401	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	1.	0.	2.	0.	0.	1.
421	0.	0.	0.	1.	1.	0.	0.	0.	0.	2.
431	0.	0.	0.	2.	0.	2.	1.	0.	0.	1.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	1.	0.	2.	0.	0.	0.	1.	0.
461	1.	1.	1.	1.	2.	4.	2.	1.	2.	4.
471	2.	7.	5.	6.	4.	5.	4.	3.	5.	3.
481	2.	0.	2.	0.	0.	0.	0.	1.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2818-DUP-PU
 File ID: 4a4355.CNF

Counted on: 12/ 4/98 @22:47
 Detector: AEA4
 Geometry number: 1
 Count time: 28808. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	970.5	970.5	360.540	360.540	10.000	2.841	5.000	0.901
2	26.4	26.4	302.270	301.921	72.000	1.000	36.000	0.100
3	46.4	46.4	286.146	286.018	10.000	1.826	5.000	0.821
4?	8.0	8.0	229.571	229.444	12.000	6.704	6.000	2.028

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs. Diff.				
1	Pu236	0.856	5.755	5.743	0.0120.01	32.13	1.6	139.9
2	Pu238	0.057	5.487	5.473	0.0140.00	2.13	6.5	12.6
	Am241		5.479	5.473	0.006			9.7
3	Th228	0.030	5.400	5.400	0.0000.01	1.13	13.3	6.8
4		????		5.140		0.32	24.3	
Totals:			0.943	<--valid peaks only-->		35.38		

DETECTOR CALIBRATION

Energy (MEV) = 4.084 + (0.0046)*Channel
 Energy range (MeV): 4.084 TO 6.440
 Efficiency = 0.2343 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18019.0	100.000
Smoothed	18018.9	99.999
Composite fit	17140.6	95.125
Residuals	878.4	4.875

Analyzed by:

SLH2

Spectrum 4a4355.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 5353.0

```

4
4
4
4
4
2
2
2
2.
3.
2.3..
32.
.2
1..
.1..
...1..
.....1.
.....1
.....1...
.....
.....1
.....1.
...

```

1

Raw Data Dump for AEA Spectrum: 4a4355.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	0.	0.	1.	2.	0.	0.	0.	0.	1.	1.
31	0.	0.	1.	0.	1.	1.	1.	0.	0.	0.
41	0.	2.	2.	0.	0.	1.	0.	0.	1.	0.
51	1.	1.	0.	0.	0.	0.	2.	1.	0.	0.
61	0.	1.	0.	0.	2.	0.	1.	0.	1.	0.
71	0.	0.	0.	0.	1.	1.	0.	1.	3.	0.
81	0.	0.	1.	0.	0.	1.	0.	1.	0.	0.
91	0.	2.	1.	1.	1.	0.	0.	1.	0.	0.
101	1.	1.	1.	2.	1.	1.	0.	0.	1.	0.
111	0.	1.	2.	1.	0.	3.	2.	2.	1.	0.
121	2.	0.	0.	1.	0.	2.	1.	1.	2.	0.
131	2.	1.	1.	1.	0.	2.	0.	2.	1.	0.
141	2.	2.	1.	0.	0.	2.	2.	1.	3.	0.
151	4.	3.	1.	3.	1.	2.	2.	5.	2.	3.
161	1.	0.	1.	2.	1.	0.	1.	0.	2.	1.
171	1.	2.	0.	0.	2.	0.	4.	4.	1.	1.
181	3.	2.	2.	0.	1.	1.	2.	2.	3.	2.
191	4.	4.	4.	3.	3.	4.	3.	1.	1.	2.
201	1.	3.	3.	3.	4.	4.	7.	6.	6.	4.
211	3.	3.	7.	3.	6.	4.	4.	2.	8.	7.
221	9.	6.	6.	11.	9.	12.	7.	8.	17.	8.
231	17.	4.	10.	6.	5.	6.	3.	11.	6.	7.
241	13.	7.	13.	5.	13.	6.	7.	10.	12.	13.
251	14.	7.	14.	13.	17.	14.	15.	11.	15.	18.
261	19.	15.	14.	23.	13.	16.	16.	28.	36.	22.
271	27.	21.	24.	33.	17.	21.	18.	31.	40.	38.
281	34.	45.	53.	57.	58.	80.	80.	58.	43.	16.
291	15.	23.	24.	24.	25.	23.	24.	19.	27.	25.
301	30.	40.	34.	33.	19.	19.	35.	28.	29.	38.
311	22.	35.	37.	29.	32.	53.	33.	37.	43.	48.
321	58.	63.	58.	73.	64.	70.	60.	63.	91.	90.
331	83.	93.	107.	126.	118.	139.	156.	149.	180.	177.
341	193.	197.	228.	247.	235.	289.	333.	408.	441.	515.
351	603.	601.	561.	504.	507.	584.	643.	761.	888.	1119.
361	1209.	1075.	762.	320.	94.	23.	1.	1.	1.	0.
371	2.	0.	0.	1.	1.	0.	2.	2.	0.	0.
381	1.	0.	0.	0.	0.	1.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	2.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	1.	0.	0.	2.	0.	0.	0.
421	0.	1.	1.	0.	0.	1.	1.	0.	0.	0.
431	0.	0.	0.	0.	0.	1.	1.	0.	1.	0.
441	0.	0.	1.	0.	0.	0.	0.	1.	1.	0.
451	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
461	0.	0.	0.	3.	1.	1.	0.	0.	1.	2.
471	2.	2.	6.	5.	4.	7.	2.	3.	2.	2.
481	0.	1.	0.	0.	2.	0.	0.	0.	0.	1.
491	0.	0.	0.	0.	1.	0.	0.	0.	1.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2819-SAM-PU
File ID: 5a5286.CNF

Counted on: 12/ 4/98 @22:49
Detector: AEA5
Geometry number: 1
Count time: 28808. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1047.3	1047.3	363.567	363.567	10.000	2.782	5.000	0.925
2	71.1	71.1	288.967	288.938	10.000	5.493	5.000	1.708

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err	d/m	Activity	
			Exp.	Obs.	Diff.					
1	Pu236	0.865	5.755	5.764	-0.0090	0.01	33.45	1.5	142.2	0.640E-04
	Cm243		5.779	5.764	0.015				190.9	0.860E-04
2	Th228	0.069	5.400	5.420	-0.0200	0.03	2.68	5.6	15.7	0.708E-05
Totals:		0.934	<--valid peaks only-->				36.13			

DETECTOR CALIBRATION

Energy (MEV) = 4.091 + (0.0046)*Channel
Energy range (MeV): 4.091 TO 6.446
Efficiency = 0.2401 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18567.0	100.000
Smoothed	18567.0	100.000
Composite fit	17349.6	93.443
Residuals	1217.4	6.557

Analyzed by: _____

SLH2

Display Max.: 6368.7

```

2.
.2
...2
...2
.
.
1.
.1..
....1.
.....1
.....1.
.....1
.....1
.....1
.....1
.....1
.....1

```


Raw Data Dump for AEA Spectrum: 5a5286.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	1.	0.	0.	0.	0.	0.
31	0.	0.	2.	0.	0.	0.	0.	1.	0.
41	1.	0.	0.	0.	2.	0.	1.	2.	0.
51	0.	0.	0.	0.	1.	0.	0.	0.	0.
61	1.	0.	1.	0.	1.	0.	2.	2.	0.
71	1.	0.	2.	1.	2.	3.	1.	0.	1.
81	0.	2.	1.	0.	1.	0.	0.	0.	0.
91	0.	0.	0.	0.	0.	2.	1.	0.	1.
101	2.	1.	1.	0.	0.	0.	0.	0.	1.
111	0.	2.	1.	2.	0.	1.	1.	0.	0.
121	2.	0.	2.	0.	1.	1.	1.	0.	1.
131	0.	0.	0.	1.	0.	0.	2.	0.	2.
141	0.	0.	1.	1.	1.	1.	1.	0.	2.
151	0.	0.	1.	0.	1.	1.	1.	1.	3.
161	1.	1.	1.	4.	1.	4.	1.	0.	1.
171	1.	0.	0.	1.	2.	1.	1.	1.	1.
181	0.	2.	1.	0.	0.	0.	3.	3.	1.
191	2.	2.	1.	3.	4.	2.	3.	4.	1.
201	0.	1.	1.	1.	2.	3.	5.	2.	3.
211	2.	4.	7.	4.	2.	4.	4.	5.	1.
221	2.	5.	8.	5.	3.	6.	6.	7.	7.
231	6.	4.	3.	4.	7.	6.	6.	4.	10.
241	7.	4.	8.	11.	4.	8.	4.	6.	6.
251	6.	11.	6.	8.	7.	11.	12.	14.	14.
261	11.	16.	16.	12.	19.	16.	22.	22.	30.
271	32.	31.	28.	28.	19.	31.	20.	19.	32.
281	21.	34.	36.	38.	59.	73.	67.	80.	72.
291	68.	39.	23.	15.	13.	17.	27.	21.	26.
301	18.	28.	15.	25.	25.	25.	15.	28.	28.
311	23.	29.	21.	33.	18.	34.	31.	34.	48.
321	40.	58.	37.	53.	53.	57.	54.	61.	66.
331	81.	82.	83.	83.	88.	99.	107.	116.	127.
341	164.	144.	185.	160.	191.	228.	227.	259.	306.
351	362.	503.	559.	604.	609.	601.	560.	550.	613.
361	817.	938.	1178.	1332.	1165.	785.	380.	97.	11.
371	2.	2.	2.	0.	1.	0.	0.	0.	0.
381	1.	0.	1.	0.	0.	0.	0.	0.	1.
391	0.	0.	0.	0.	1.	0.	0.	1.	0.
401	0.	0.	0.	0.	1.	0.	0.	0.	1.
411	0.	0.	1.	0.	1.	1.	0.	1.	0.
421	1.	0.	0.	0.	1.	1.	0.	0.	0.
431	1.	0.	1.	1.	0.	2.	1.	0.	1.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	1.	1.	1.	0.	1.	1.	0.	0.
461	1.	0.	0.	1.	2.	1.	2.	2.	1.
471	1.	4.	7.	5.	1.	4.	6.	10.	9.
481	2.	4.	3.	3.	3.	1.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2819-DUP-PU
File ID: 6a6291.CNF

Counted on: 12/ 4/98 @22:49
Detector: AEA6
Geometry number: 1
Count time: 28807. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1123.8	1123.8	360.824	360.824	10.000	2.625	5.000	0.985
2	73.4	73.4	286.184	286.174	12.000	4.187	6.000	2.726
3	20.9	20.9	268.728	268.109	16.000	5.380	8.000	1.382

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid Exp.	Obs.	Diff.	FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
1	Pu236	0.871	5.755	5.749	0.0060	0.01	32.78	1.6	149.3	0.672E-04
2	Th228	0.048	5.400	5.405	-0.0050	0.02	1.80	6.7	11.3	0.509E-05
3		0.023		5.322		0.02	0.88	12.2	3.9	0.178E-05

Totals: 0.942 <--valid peaks only--> 35.46

DETECTOR CALIBRATION

Energy (MEV) = 4.089 + (0.0046)*Channel
Energy range (MeV): 4.089 TO 6.444
Efficiency = 0.2241 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18079.0	100.000
Smoothed	18079.0	100.000
Composite fit	17026.2	94.177
Residuals	1052.8	5.823

Analyzed by:

SLH2

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

5748.5

```

3
3
3
3.
3.
....2
.2
.
..
1.
.1..
.....1.
.....1
.....1...
.....1.....
.....1.....
.....1.
...

```

Raw Data Dump for AEA Spectrum: 6a6291.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	1.	0.	1.	0.	0.	0.	2.	0.	0.
31	1.	0.	0.	1.	0.	0.	0.	0.	0.	2.
41	0.	1.	0.	0.	0.	1.	1.	1.	0.	0.
51	0.	0.	0.	2.	0.	2.	0.	1.	1.	1.
61	0.	0.	2.	0.	0.	0.	0.	2.	0.	1.
71	1.	1.	0.	0.	0.	1.	1.	2.	2.	0.
81	2.	1.	2.	0.	0.	1.	0.	0.	0.	0.
91	1.	1.	1.	0.	2.	0.	1.	0.	1.	0.
101	0.	0.	1.	0.	0.	2.	0.	2.	1.	1.
111	2.	0.	1.	0.	0.	2.	1.	0.	3.	1.
121	0.	1.	1.	1.	1.	2.	2.	2.	0.	1.
131	2.	0.	2.	0.	0.	0.	2.	2.	1.	2.
141	1.	2.	0.	0.	2.	0.	0.	3.	2.	1.
151	1.	1.	2.	0.	0.	2.	2.	0.	1.	2.
161	0.	1.	1.	1.	0.	3.	1.	0.	1.	1.
171	1.	1.	1.	2.	2.	1.	3.	0.	3.	1.
181	1.	3.	0.	1.	1.	2.	1.	1.	2.	2.
191	2.	2.	5.	3.	2.	2.	4.	3.	2.	2.
201	3.	2.	2.	2.	5.	0.	0.	2.	1.	2.
211	4.	4.	1.	2.	4.	5.	2.	7.	1.	3.
221	1.	3.	8.	8.	6.	5.	8.	5.	9.	6.
231	2.	1.	6.	3.	2.	4.	2.	4.	3.	8.
241	2.	10.	6.	8.	7.	3.	4.	1.	11.	10.
251	6.	10.	7.	8.	12.	9.	3.	9.	11.	8.
261	8.	16.	19.	20.	21.	21.	22.	29.	29.	26.
271	17.	21.	22.	25.	16.	17.	18.	20.	29.	28.
281	37.	34.	48.	69.	75.	97.	87.	66.	51.	24.
291	9.	14.	11.	13.	13.	20.	15.	18.	19.	17.
301	19.	20.	30.	20.	28.	21.	18.	23.	24.	24.
311	29.	23.	24.	28.	23.	27.	39.	35.	38.	28.
321	33.	42.	45.	42.	56.	50.	56.	53.	49.	68.
331	69.	87.	85.	81.	85.	94.	120.	128.	110.	123.
341	156.	154.	179.	191.	231.	308.	370.	456.	515.	
351	647.	647.	610.	569.	541.	545.	673.	812.	986.	1245.
361	1386.	1356.	973.	438.	109.	21.	3.	1.	1.	0.
371	0.	1.	0.	0.	0.	0.	0.	1.	1.	0.
381	0.	0.	1.	2.	0.	0.	0.	0.	0.	0.
391	1.	0.	1.	2.	0.	2.	0.	0.	0.	0.
401	0.	0.	1.	1.	0.	0.	0.	0.	1.	0.
411	0.	2.	0.	0.	0.	0.	0.	1.	0.	0.
421	0.	1.	0.	1.	0.	0.	1.	0.	1.	2.
431	0.	1.	1.	0.	2.	0.	1.	0.	0.	1.
441	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	1.	0.	1.	1.	1.	0.	2.	0.	0.
461	0.	0.	0.	1.	0.	1.	1.	2.	4.	2.
471	2.	5.	4.	3.	8.	4.	8.	6.	7.	2.
481	2.	1.	0.	1.	0.	1.	0.	0.	0.	0.
491	0.	1.	0.	1.	0.	0.	0.	0.	1.	0.
511	0.	0.								

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2820-SAM-PU
 File ID: 7a7319.CNF

Counted on: 12/ 4/98 @22:50
 Detector: AEA7
 Geometry number: 1
 Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1128.6	1128.6	363.524	363.524	8.000	2.519	4.000	0.901
2	78.0	78.0	288.938	288.928	10.000	3.859	5.000	1.686
3	10.5	10.5	255.505	255.197	26.000	89.357	13.000	12.679

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count	%err	Activity		
		Exp.	Obs.	Diff.	FWHM	Rate	c/m	@95	d/m	uCi/ea
1	Pu236	0.871	5.755	5.764	-.0090.01	33.77	1.5	143.0	0.644E-04	
	Cm243		5.779	5.764	0.015			191.9	0.865E-04	
2	Th228	0.058	5.400	5.420	-.0200.02	2.24	6.0	13.1	0.591E-05	
3	Pu239	0.031	5.147	5.265	-.1180.41	1.18	14.3	4.9	0.221E-05	
	Am243		5.266	5.265	0.001			5.0	0.223E-05	
Totals:			0.960	<--valid peaks only-->			37.19			

DETECTOR CALIBRATION

Energy (MEV) = 4.091 + (0.0046)*Channel
 Energy range (MeV): 4.091 TO 6.447
 Efficiency = 0.2410 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18608.0	100.000
Smoothed	18607.5	99.998
Composite fit	17855.5	95.956
Residuals	752.5	4.044

Analyzed by:

SLH2

3

```

3
3
3
3
3
3
3
3
3
3
3
3..2
...2
.
.
.
.1.
...1..
.....1
.....1.
.....1
.....1.....1
.....1.....
```

Raw Data Dump for AEA Spectrum: 7a7319.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	1.	1.
21	0.	0.	0.	0.	1.	1.	0.	0.	0.
31	0.	0.	0.	0.	1.	1.	0.	1.	1.
41	0.	0.	0.	0.	0.	2.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	0.	1.
61	0.	0.	0.	1.	1.	2.	0.	0.	0.
71	0.	1.	0.	0.	2.	0.	0.	1.	4.
81	1.	0.	0.	0.	0.	0.	1.	0.	0.
91	0.	0.	0.	0.	0.	1.	2.	0.	1.
101	0.	0.	0.	1.	0.	0.	2.	1.	0.
111	1.	1.	0.	1.	1.	0.	1.	1.	0.
121	0.	1.	1.	0.	1.	0.	1.	0.	1.
131	0.	0.	1.	1.	1.	0.	0.	0.	1.
141	0.	0.	0.	0.	0.	2.	2.	3.	1.
151	1.	2.	0.	2.	2.	0.	0.	1.	0.
161	1.	0.	1.	1.	0.	2.	1.	0.	1.
171	1.	1.	2.	0.	1.	1.	2.	0.	1.
181	0.	0.	3.	1.	1.	1.	3.	1.	2.
191	2.	3.	0.	3.	2.	3.	1.	4.	1.
201	0.	0.	0.	1.	2.	2.	2.	0.	3.
211	2.	2.	2.	5.	2.	1.	4.	4.	11.
221	6.	6.	5.	6.	4.	4.	4.	5.	5.
231	6.	5.	6.	5.	3.	5.	3.	6.	4.
241	5.	3.	5.	7.	6.	8.	5.	6.	4.
251	11.	7.	6.	7.	15.	12.	8.	7.	8.
261	15.	14.	12.	15.	18.	14.	21.	30.	28.
271	28.	34.	27.	25.	14.	32.	15.	19.	19.
281	25.	32.	34.	38.	45.	44.	64.	88.	103.
291	60.	29.	22.	17.	18.	16.	18.	15.	13.
301	12.	23.	27.	26.	26.	22.	28.	22.	23.
311	23.	28.	26.	22.	20.	42.	30.	25.	34.
321	29.	45.	41.	41.	47.	58.	46.	48.	70.
331	71.	89.	77.	76.	86.	85.	93.	102.	128.
341	139.	124.	142.	209.	220.	216.	227.	269.	275.
351	395.	533.	572.	668.	637.	611.	570.	486.	562.
361	830.	1074.	1252.	1449.	1331.	727.	299.	57.	12.
371	1.	2.	0.	2.	0.	0.	1.	0.	1.
381	0.	1.	0.	0.	0.	0.	0.	0.	2.
391	0.	0.	0.	0.	1.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	1.	2.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	0.	1.	0.	1.	1.	0.	0.	1.	2.
431	1.	1.	1.	2.	0.	3.	2.	0.	0.
441	0.	0.	0.	0.	0.	1.	1.	0.	0.
451	1.	0.	1.	1.	0.	0.	1.	0.	0.
461	1.	0.	2.	0.	0.	1.	0.	1.	3.
471	0.	0.	3.	1.	3.	5.	7.	5.	6.
481	4.	4.	3.	5.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	1.	0.	0.
511	1.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2820-DUP-PU
File ID: 8a8297.CNF

Counted on: 12/ 4/98 @22:51
Detector: AEA8
Geometry number: 1
Count time: 28808. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1063.2	1063.2	362.856	362.856	10.000	2.437	5.000	0.896
2	71.0	71.0	288.139	288.128	12.000	7.283	6.000	2.408
3	11.4	11.4	272.357	270.305	20.000	1.070	10.000	0.139

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count	%err	Activity	
		Exp.	Obs.	Diff.	FWHM	Rate	c/m	uCi/ea	
1	Pu236	0.839	5.755	5.759	-0.0040	0.01	31.06	1.6	140.3
	Cm243		5.779	5.759	0.020				188.4
2	Th228	0.074	5.400	5.415	-0.0150	0.03	2.75	5.5	17.1
3		0.019		5.333	0.00		0.72	14.7	3.2
Totals:		0.933	<--valid peaks only-->				34.53		

DETECTOR CALIBRATION

Energy(MEV) = 4.090 + (0.0046)*Channel
Energy range (MeV): 4.090 TO 6.445
Efficiency = 0.2259 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17771.0	100.000
Smoothed	17771.0	100.000
Composite fit	16579.1	93.293
Residuals	1191.9	6.707

Analyzed by:

SLH2

5889.6

[illegible]

Raw Data Dump for AEA Spectrum: 8a8297.CNF

HNF-1663 REV.0

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	0.	0.	0.	0.	0.	0.	0.	0.	1.
31	0.	0.	0.	1.	0.	0.	0.	1.	0.
41	0.	0.	0.	0.	0.	2.	0.	0.	0.
51	2.	0.	2.	0.	0.	0.	0.	0.	1.
61	0.	0.	0.	1.	0.	0.	0.	2.	0.
71	0.	0.	0.	0.	0.	0.	1.	0.	0.
81	0.	1.	0.	1.	1.	1.	0.	0.	1.
91	2.	2.	3.	0.	1.	2.	2.	0.	2.
101	0.	2.	0.	1.	0.	1.	1.	0.	0.
111	0.	0.	0.	0.	0.	1.	0.	4.	0.
121	2.	0.	2.	1.	1.	1.	0.	1.	1.
131	2.	2.	0.	3.	2.	0.	0.	1.	1.
141	1.	2.	4.	0.	0.	1.	2.	0.	2.
151	0.	0.	1.	4.	1.	0.	3.	1.	0.
161	3.	3.	1.	1.	0.	2.	2.	1.	0.
171	2.	1.	3.	1.	1.	0.	0.	4.	0.
181	1.	3.	4.	4.	2.	2.	3.	2.	3.
191	2.	2.	3.	0.	3.	2.	2.	2.	2.
201	2.	3.	2.	3.	1.	1.	5.	2.	3.
211	3.	4.	3.	3.	0.	6.	3.	6.	1.
221	3.	4.	1.	4.	4.	6.	3.	8.	3.
231	4.	6.	3.	3.	7.	2.	8.	5.	8.
241	3.	6.	2.	5.	7.	7.	7.	8.	7.
251	11.	11.	11.	16.	7.	13.	12.	9.	12.
261	15.	18.	11.	32.	20.	13.	12.	24.	26.
271	24.	34.	28.	26.	16.	22.	22.	24.	28.
281	34.	29.	36.	47.	58.	62.	77.	83.	94.
291	44.	29.	20.	20.	26.	15.	24.	21.	20.
301	17.	31.	27.	19.	30.	34.	27.	28.	18.
311	27.	30.	28.	35.	31.	32.	31.	46.	42.
321	46.	50.	71.	44.	61.	72.	37.	52.	74.
331	88.	73.	80.	96.	83.	93.	101.	118.	132.
341	144.	139.	169.	155.	193.	206.	227.	261.	296.
351	445.	496.	649.	634.	538.	493.	491.	452.	621.
361	904.	1129.	1367.	1285.	901.	357.	94.	16.	1.
371	2.	2.	0.	0.	1.	0.	0.	1.	0.
381	1.	0.	1.	0.	0.	1.	1.	0.	2.
391	1.	0.	0.	0.	1.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	1.	0.	0.	0.	0.	0.	0.
421	0.	0.	2.	1.	0.	1.	1.	0.	0.
431	2.	1.	1.	0.	2.	0.	0.	0.	0.
441	0.	0.	0.	1.	0.	0.	0.	0.	0.
451	1.	0.	0.	0.	0.	3.	0.	1.	0.
461	0.	1.	0.	1.	0.	2.	2.	1.	1.
471	1.	3.	5.	1.	3.	6.	8.	5.	7.
481	2.	1.	2.	1.	1.	0.	0.	1.	1.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							1.

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2821-SAM-PU
File ID: 9a9210.CNF

Counted on: 12/ 4/98 @22:52
Detector: AEA9
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1017.4	1017.4	361.191	361.191	10.000	2.630	5.000	0.873
2	73.4	73.4	286.264	286.245	12.000	6.028	6.000	2.886
3	17.9	17.9	269.989	268.936	10.000	3.480	5.000	0.688

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Exp.	Centroid Obs.	Diff.	FWHM	Count Rate	%err c/m	@95	d/m	Activity uCi/ea	
1	Pu236	0.849	5.755	5.759	-0.0040	0.01	32.25	1.6		138.4	0.623E-04	
2	Th228	0.060	5.400	5.414	-0.0140	0.03	2.27	6.1		13.4	0.605E-05	
3		0.023		5.334		0.02	0.86	12.8		3.6	0.163E-05	

Totals:		0.932	<--valid peaks only-->				35.38					

DETECTOR CALIBRATION

Energy (MEV) = 4.097 + (0.0046)*Channel
Energy range (MeV): 4.097 TO 6.453
Efficiency = 0.2378 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18230.0	100.000
Smoothed	18229.9	100.000
Composite fit	16983.9	93.164
Residuals	1246.1	6.836

Analyzed by:

SLH2

Spectrum 9a9210.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5093.3

```

3
3
3
3
3.
32.
.....2
..2
..
...
1...
...1..
.....1..
.....1.
.....1....
.....1.....
.....1.....
.....1
.....

```

Raw Data Dump for AEA Spectrum: 9a9210.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	0.	2.	0.	0.	0.	0.	0.	2.	0.	0.
31	0.	1.	1.	0.	1.	0.	0.	0.	0.	0.
41	0.	0.	0.	2.	0.	1.	0.	2.	0.	1.
51	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
61	0.	2.	0.	0.	0.	0.	0.	1.	2.	0.
71	2.	1.	0.	1.	2.	0.	0.	0.	0.	2.
81	0.	1.	1.	0.	0.	0.	4.	0.	1.	0.
91	0.	2.	0.	0.	0.	0.	2.	1.	0.	4.
101	0.	0.	0.	1.	0.	0.	1.	1.	0.	0.
111	0.	0.	0.	0.	2.	1.	0.	1.	2.	1.
121	1.	1.	1.	0.	3.	0.	1.	2.	0.	2.
131	1.	0.	1.	0.	3.	1.	1.	1.	2.	3.
141	0.	1.	2.	2.	2.	2.	1.	2.	2.	1.
151	0.	1.	0.	0.	1.	0.	0.	2.	0.	1.
161	0.	0.	0.	1.	1.	2.	0.	3.	1.	0.
171	0.	1.	0.	1.	0.	1.	0.	4.	1.	0.
181	2.	0.	9.	3.	1.	3.	2.	2.	3.	4.
191	0.	2.	0.	2.	4.	0.	4.	2.	5.	2.
201	2.	1.	2.	3.	4.	3.	5.	2.	4.	5.
211	3.	0.	4.	9.	8.	1.	3.	0.	6.	2.
221	9.	9.	4.	8.	5.	6.	5.	7.	5.	5.
231	3.	4.	8.	4.	6.	3.	3.	5.	6.	8.
241	6.	11.	8.	9.	10.	5.	10.	5.	6.	6.
251	8.	6.	12.	10.	10.	13.	10.	17.	12.	18.
261	19.	14.	14.	22.	16.	24.	22.	24.	33.	31.
271	30.	23.	23.	22.	22.	28.	33.	35.	30.	34.
281	37.	41.	47.	64.	75.	99.	83.	73.	47.	34.
291	26.	12.	22.	18.	22.	17.	27.	25.	18.	19.
301	25.	28.	35.	31.	23.	31.	29.	28.	32.	43.
311	32.	28.	42.	31.	44.	37.	50.	47.	47.	46.
321	52.	47.	60.	56.	69.	66.	72.	80.	90.	103.
331	86.	103.	95.	115.	121.	110.	131.	145.	151.	155.
341	194.	206.	211.	241.	248.	273.	261.	356.	429.	496.
351	545.	673.	616.	520.	518.	479.	549.	672.	808.	1022.
361	1272.	1260.	1020.	549.	201.	40.	7.	2.	0.	0.
371	2.	1.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.	2.
421	1.	1.	1.	1.	0.	3.	1.	0.	0.	0.
431	1.	1.	0.	0.	0.	0.	1.	0.	1.	1.
441	0.	1.	1.	0.	0.	1.	0.	0.	0.	0.
451	0.	1.	0.	0.	0.	0.	0.	1.	1.	0.
461	3.	1.	0.	1.	1.	1.	3.	0.	3.	2.
471	4.	5.	4.	3.	6.	10.	11.	2.	2.	3.
481	6.	3.	3.	1.	0.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2821-DUP-PU
 File ID: 10a1049.CNF

Counted on: 12/ 4/98 @22:53
 Detector: AEA10
 Geometry number: 1
 Count time: 28806. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1010.0	1010.0	362.026	362.026	10.000	2.858	5.000	0.914
2	67.5	67.5	287.803	287.771	10.000	6.200	5.000	2.370
3	14.4	14.4	269.429	268.328	12.000	3.193	6.000	0.443

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea	
			Exp.	Obs. Diff.						
1	Pu236	0.861	5.755	5.747	0.0080	0.1	33.27	1.6	143.6	0.647E-04
2	Th228	0.060	5.400	5.405	-0.0050	0.3	2.33	6.0	13.9	0.626E-05
3		0.024		5.316	0.01	0.91	12.2		3.9	0.174E-05
Totals:			0.945	<--valid peaks only-->			36.51			

DETECTOR CALIBRATION

Energy(MEV) = 4.082 + (0.0046)*Channel
 Energy range (MeV): 4.082 TO 6.437
 Efficiency = 0.2364 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18542.0	100.000
Smoothed	18542.0	100.000
Composite fit	17529.9	94.542
Residuals	1012.1	5.458

Analyzed by:

SLH2

Spectrum 10a1049.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,.., etc   Display Max.: 5330.5

```

3
3
3
3
3
3
3
3.
.2
....2
...2
..
...
.1..
...1.
.....1.
.....1
.....1..
.....1
.....1
.....1
.....1
.....1

```

Raw Data Dump for AEA Spectrum: 10a1049.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	0.	0.	0.
21	0.	0.	0.	2.	0.	0.	2.	0.	0.
31	0.	0.	0.	0.	0.	1.	0.	0.	1.
41	1.	0.	2.	0.	1.	1.	1.	0.	0.
51	0.	1.	0.	1.	3.	1.	0.	0.	0.
61	0.	2.	0.	1.	0.	1.	0.	0.	0.
71	1.	1.	2.	0.	1.	0.	0.	0.	1.
81	2.	0.	0.	0.	1.	0.	1.	2.	0.
91	0.	3.	1.	0.	0.	0.	0.	1.	2.
101	0.	1.	1.	0.	1.	0.	2.	2.	0.
111	1.	0.	0.	3.	1.	1.	2.	1.	1.
121	0.	1.	1.	1.	1.	2.	1.	1.	0.
131	1.	1.	2.	3.	0.	3.	0.	0.	0.
141	1.	1.	2.	1.	3.	2.	1.	2.	0.
151	2.	1.	1.	1.	3.	1.	1.	0.	3.
161	4.	1.	5.	2.	3.	2.	2.	0.	1.
171	2.	0.	3.	1.	2.	2.	4.	2.	1.
181	3.	3.	1.	1.	2.	3.	5.	4.	3.
191	4.	5.	1.	2.	1.	3.	5.	4.	1.
201	1.	2.	2.	1.	2.	7.	3.	1.	6.
211	1.	1.	7.	1.	6.	5.	8.	2.	3.
221	1.	8.	5.	4.	6.	5.	4.	6.	2.
231	6.	7.	2.	3.	9.	4.	6.	6.	9.
241	6.	4.	3.	9.	3.	7.	10.	12.	9.
251	8.	11.	8.	5.	10.	22.	18.	9.	17.
261	21.	18.	13.	12.	21.	15.	23.	30.	27.
271	21.	26.	17.	19.	23.	30.	25.	25.	33.
281	34.	45.	34.	58.	64.	56.	67.	85.	84.
291	37.	13.	20.	16.	25.	25.	28.	27.	15.
301	24.	35.	37.	30.	25.	34.	29.	36.	34.
311	23.	25.	40.	31.	48.	41.	30.	47.	30.
321	48.	51.	53.	56.	61.	45.	81.	74.	78.
331	96.	93.	82.	112.	113.	128.	121.	143.	142.
341	134.	163.	188.	199.	245.	254.	260.	353.	341.
351	495.	613.	632.	630.	551.	540.	559.	616.	749.
361	1039.	1210.	1214.	976.	558.	196.	37.	3.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	1.	0.	1.	1.	0.	0.
411	0.	0.	0.	1.	0.	0.	0.	0.	0.
421	0.	0.	0.	0.	1.	1.	1.	0.	2.
431	0.	1.	0.	2.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	1.	0.
451	0.	0.	0.	0.	0.	0.	1.	0.	1.
461	2.	1.	1.	0.	1.	2.	0.	0.	0.
471	2.	2.	1.	6.	9.	5.	7.	4.	5.
481	4.	4.	4.	0.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2939-SAM-PU
File ID: 11a1146.CNF

Counted on: 12/ 4/98 @22:54
Detector: AEA11
Geometry number: 1
Count time: 28800. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	326.4	326.4	353.885	353.885	30.000	28.773	15.000	2.786
2	11.9	11.9	232.613	231.523	38.000	1.000	19.000	0.100
3?	3.4	3.4	184.086	183.660	8.000	1.102	4.000	0.204

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m	Activity d/m	Activity uCi/ea
		Exp.	Obs.	Diff.	FWHM				
1	Ra224	0.917	5.686	5.702	-0.016	0.13	35.09 1.5	156.7	0.706E-04
	Pu236		5.755	5.702	0.053			150.3	0.677E-04
2	Pu239	0.025	5.147	5.139	0.008	0.00	0.96 13.7	4.0	0.181E-05
	Pu240		5.144	5.139	0.005			4.0	0.181E-05
3		????		4.919			0.16 49.8		
Totals:			0.942	<--valid peaks only-->			36.05		

DETECTOR CALIBRATION

Energy (MEV) = 4.074 + (0.0046)*Channel
Energy range (MeV): 4.074 TO 6.429
Efficiency = 0.2382 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18366.0	100.000
Smoothed	18366.0	100.000
Composite fit	17379.1	94.626
Residuals	986.9	5.374

Analyzed by:

SLH2

Raw Data Dump for AEA Spectrum: 11a1146.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	2.	4.	1.	1.	1.	1.
21	2.	0.	3.	1.	1.	2.	3.	0.	1.	0.
31	1.	4.	1.	0.	2.	4.	3.	2.	2.	1.
41	2.	3.	3.	3.	1.	3.	5.	3.	3.	0.
51	1.	1.	1.	2.	3.	1.	2.	2.	5.	1.
61	0.	4.	1.	7.	0.	3.	1.	4.	5.	3.
71	2.	2.	5.	6.	2.	2.	3.	4.	2.	4.
81	3.	4.	2.	3.	1.	3.	4.	3.	2.	1.
91	1.	0.	5.	3.	4.	3.	3.	4.	5.	3.
101	1.	0.	4.	3.	5.	5.	3.	0.	2.	0.
111	5.	4.	4.	5.	0.	4.	3.	6.	3.	2.
121	4.	0.	3.	5.	0.	2.	4.	4.	6.	4.
131	2.	8.	6.	5.	7.	5.	6.	3.	4.	5.
141	2.	7.	7.	1.	2.	4.	8.	7.	6.	5.
151	5.	6.	5.	3.	9.	4.	9.	3.	8.	5.
161	7.	8.	5.	5.	3.	6.	15.	8.	2.	5.
171	6.	7.	11.	9.	8.	9.	5.	8.	6.	10.
181	8.	10.	10.	11.	12.	12.	3.	9.	6.	6.
191	9.	10.	11.	9.	7.	14.	8.	9.	15.	7.
201	10.	14.	10.	6.	9.	15.	10.	10.	8.	13.
211	16.	14.	17.	18.	20.	18.	11.	18.	25.	11.
221	15.	19.	28.	26.	19.	21.	27.	17.	27.	20.
231	27.	25.	29.	20.	26.	28.	12.	23.	29.	22.
241	21.	17.	28.	24.	27.	39.	27.	32.	31.	25.
251	33.	37.	44.	49.	22.	24.	40.	41.	40.	51.
261	34.	54.	41.	54.	40.	54.	44.	42.	51.	50.
271	46.	47.	56.	46.	55.	45.	41.	64.	46.	58.
281	72.	60.	65.	70.	62.	72.	80.	68.	63.	69.
291	66.	84.	69.	78.	57.	67.	75.	76.	72.	81.
301	75.	82.	72.	76.	77.	88.	92.	89.	125.	104.
311	104.	102.	107.	106.	138.	121.	140.	128.	151.	148.
321	142.	139.	150.	167.	168.	181.	170.	161.	180.	188.
331	187.	189.	186.	210.	202.	227.	230.	247.	250.	270.
341	252.	269.	285.	255.	298.	308.	315.	292.	282.	342.
351	310.	337.	324.	351.	306.	318.	347.	297.	309.	277.
361	240.	239.	243.	247.	252.	231.	181.	85.	25.	2.
371	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	1.	0.	1.	0.	0.
391	0.	0.	1.	0.	0.	0.	0.	0.	1.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	1.	1.	1.	0.	0.
421	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
431	0.	0.	3.	0.	0.	0.	0.	0.	0.	1.
441	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	1.	0.	0.	0.	0.	1.	0.	1.
461	1.	0.	1.	0.	1.	0.	1.	0.	0.	0.
471	0.	0.	1.	1.	1.	0.	0.	2.	2.	0.
481	1.	2.	0.	2.	2.	2.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2939-DUP-PU
File ID: 12a1220.CNF

Counted on: 12/ 4/98 @22:55
Detector: AEA12
Geometry number: 1
Count time: 28805. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	801.2	801.2	359.885	359.885	10.000	2.895	5.000	0.640
2?	16.4	16.4	300.883	300.883	12.000	0.000	6.000	4.895
3	41.5	41.5	285.328	285.150	12.000	5.036	6.000	1.426
4	12.4	12.4	268.729	267.394	10.000	2.589	5.000	0.517

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err	d/m	Activity
			Exp.	Obs. Diff.	c/m	@95		uCi/ea
1	Pu236	0.950	5.755	5.757 -0.0020.01	34.67	1.5	155.0	0.698E-04
2		????		5.485	0.00	1000.		
3	Th228	0.045	5.400	5.413 -.0130.02	1.63	8.4	10.0	0.453E-05
4		0.016		5.331 0.01	0.57	18.5	2.5	0.113E-05
Totals:		1.010	<--valid peaks only-->		36.87			

DETECTOR CALIBRATION

Energy(MEV) = 4.101 + (0.0046)*Channel
Energy range (MeV): 4.101 TO 6.456
Efficiency = 0.2283 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17525.0	100.000
Smoothed	17525.0	100.000
Composite fit	17700.0	100.999
Residuals	-175.0	-0.999

Analyzed by: _____

SLH2

Spectrum 12a1220.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 5113.1

```

4
4
4
4
4.
.3.
...3
3.
.1
..1
.....1
..... 1
..... 1
..... 1
.....1...
..... 1
.....1.
..

```

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
31	2.	0.	0.	1.	0.	1.	0.	0.	0.	0.
41	1.	0.	0.	0.	0.	1.	0.	0.	1.	0.
51	0.	0.	2.	0.	0.	2.	0.	0.	0.	0.
61	1.	0.	0.	0.	0.	1.	0.	0.	0.	2.
71	0.	0.	0.	0.	0.	0.	1.	1.	0.	1.
81	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
91	0.	1.	0.	1.	0.	0.	1.	0.	0.	1.
101	0.	0.	0.	0.	0.	1.	0.	2.	0.	1.
111	0.	0.	1.	0.	1.	0.	0.	1.	0.	1.
121	0.	0.	0.	0.	0.	0.	0.	1.	1.	0.
131	1.	0.	0.	1.	2.	0.	0.	0.	1.	1.
141	0.	3.	2.	3.	2.	1.	0.	1.	2.	0.
151	1.	1.	2.	1.	0.	1.	1.	1.	1.	0.
161	0.	0.	1.	2.	1.	1.	1.	2.	0.	2.
171	2.	3.	0.	0.	2.	1.	1.	0.	1.	2.
181	1.	0.	0.	1.	2.	0.	0.	0.	2.	3.
191	0.	1.	1.	1.	3.	3.	0.	0.	1.	3.
201	1.	1.	3.	0.	5.	5.	1.	2.	5.	1.
211	2.	4.	1.	2.	7.	2.	2.	3.	3.	5.
221	5.	2.	6.	6.	1.	5.	6.	4.	7.	6.
231	0.	3.	7.	7.	7.	3.	7.	6.	3.	4.
241	6.	9.	3.	10.	9.	9.	6.	5.	9.	7.
251	12.	15.	7.	7.	16.	11.	15.	12.	20.	17.
261	20.	19.	23.	24.	23.	23.	26.	30.	30.	28.
271	25.	21.	16.	26.	27.	26.	24.	30.	38.	36.
281	38.	51.	37.	47.	61.	53.	52.	37.	30.	13.
291	25.	20.	26.	20.	19.	24.	25.	27.	28.	25.
301	29.	30.	27.	22.	17.	26.	33.	30.	50.	32.
311	42.	38.	47.	53.	44.	52.	62.	63.	69.	52.
321	56.	77.	76.	66.	81.	86.	100.	91.	102.	140.
331	112.	122.	123.	122.	150.	184.	175.	204.	206.	257.
341	243.	243.	323.	274.	355.	353.	372.	462.	481.	603.
351	648.	585.	595.	536.	500.	573.	670.	656.	829.	925.
361	955.	781.	369.	122.	23.	2.	0.	0.	0.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	1.	0.	0.	0.	1.	0.	0.	1.	0.	0.
421	0.	1.	0.	0.	0.	1.	0.	0.	0.	0.
431	0.	0.	0.	0.	0.	2.	0.	1.	1.	1.
441	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
451	1.	0.	1.	0.	0.	0.	1.	0.	1.	0.
461	0.	0.	0.	1.	0.	2.	1.	0.	3.	2.
471	4.	3.	2.	6.	3.	6.	6.	2.	2.	2.
481	3.	0.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

HNF-1663 REV. 0

LABCORE Completed Worklist Report for Worklist# 26723

Analyst: crj Instrument: AB20 Book#: _____

Method: LA-953-104 Rev/Mod _____

Worklist Comment: BY-105 FOR @PU23901 RTS

Seq Type	Sample# R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@PU23901 PU23901 LIQUID		1.12E-04	1.24E-4	110.714 % Recovery	
1 STD	0	@PU23901 PU23901T LIQUID		100	7.83E+01	78.300 % Recovery	
1 STD	0	@PU23901 PU23901E LIQUID		1.00	2.02E+00	2.020 % Ct Error	
2 BLNK	0	@PU23901 PU23901 LIQUID		1	<8.17E-6	uCi/mL	
2 BLNK	0	@PU23901 PU23901T LIQUID		100	7.80E+01	78.000 % Recovery	
2 BLNK	0	@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000 uCi/mL	
3 SAMPLE	S98T002960 0	@PU23901 PU23901 LIQUID		N/A	1.15E-05	1.15E-005 uCi/mL	
3 SAMPLE	S98T002960 0	@PU23901 PU23901T LIQUID		N/A	5.51E+01	% Recovery	
3 SAMPLE	S98T002960 0	@PU23901 PU23901E LIQUID		N/A	8.38E+00	% Ct. Error	
4 DUP	S98T002960 0	@PU23901 PU23901 LIQUID		<1.15E-5	9.72E-6	RPD	
4 DUP	S98T002960 0	@PU23901 PU23901T LIQUID		100	8.58E+01	85.800 % Recovery	
4 DUP	S98T002960 0	@PU23901 PU23901E LIQUID		1.00	6.90E+00	6.900 % Ct Error	
5 SAMPLE	S98T002961 0	@PU23901 PU23901 LIQUID		N/A	7.98E-06	7.98E-006 uCi/mL	
5 SAMPLE	S98T002961 0	@PU23901 PU23901T LIQUID		N/A	7.95E+01	% Recovery	
5 SAMPLE	S98T002961 0	@PU23901 PU23901E LIQUID		N/A	8.01E+00	% Ct. Error	
6 DUP	S98T002961 0	@PU23901 PU23901 LIQUID		<7.98E-6	8.93E-6	RPD	
6 DUP	S98T002961 0	@PU23901 PU23901T LIQUID		100	7.85E+01	78.500 % Recovery	
6 DUP	S98T002961 0	@PU23901 PU23901E LIQUID		1.00	7.29E+00	7.290 % Ct Error	
7 SAMPLE	S98T002962 0	@PU23901 PU23901 LIQUID		N/A	1.11E-05	7.93E-006 uCi/mL	
7 SAMPLE	S98T002962 0	@PU23901 PU23901T LIQUID		N/A	8.10E+01	% Recovery	
7 SAMPLE	S98T002962 0	@PU23901 PU23901E LIQUID		N/A	6.48E+00	% Ct. Error	
8 DUP	S98T002962 0	@PU23901 PU23901 LIQUID		1.11E-5	1.14E-5	2.667 RPD	
8 DUP	S98T002962 0	@PU23901 PU23901T LIQUID		100	8.19E+01	81.900 % Recovery	
8 DUP	S98T002962 0	@PU23901 PU23901E LIQUID		1.00	6.49E+00	6.490 % Ct Error	
9 SAMPLE	S98T002963 0	@PU23901 PU23901 LIQUID		N/A	7.90E-06	7.90E-006 uCi/mL	
9 SAMPLE	S98T002963 0	@PU23901 PU23901T LIQUID		N/A	8.05E+01	% Recovery	
9 SAMPLE	S98T002963 0	@PU23901 PU23901E LIQUID		N/A	9.08E+00	% Ct. Error	
10 DUP	S98T002963 0	@PU23901 PU23901 LIQUID		<7.90E-6	<7.84E-6	RPD	
10 DUP	S98T002963 0	@PU23901 PU23901T LIQUID		100	7.99E+01	79.900 % Recovery	
10 DUP	S98T002963 0	@PU23901 PU23901E LIQUID		1.00	9.31E+00	9.310 % Ct Error	
11 SAMPLE	S98T002964 0	@PU23901 PU23901 LIQUID		N/A	8.13E-06	8.02E-006 uCi/mL	
11 SAMPLE	S98T002964 0	@PU23901 PU23901T LIQUID		N/A	8.00E+01	% Recovery	
11 SAMPLE	S98T002964 0	@PU23901 PU23901E LIQUID		N/A	7.73E+00	% Ct. Error	
12 DUP	S98T002964 0	@PU23901 PU23901 LIQUID		8.13E-6	<8.01E-6	RPD	
12 DUP	S98T002964 0	@PU23901 PU23901T LIQUID		100	8.04E+01	80.400 % Recovery	
12 DUP	S98T002964 0	@PU23901 PU23901E LIQUID		1.00	7.80E+00	7.800 % Ct Error	

Final page for worklist# 26723

LABCORE Completed Worklist Report for Worklist# 26723

Seq	Type	Sample#	R	A	Test	Matrix	Actual	Found	DL or Yield	Unit
Analyst Signature		Date					Analyst Signature		Date	

<u><i>Daphy Relyea</i></u>	<u><i>20 Nov 98</i></u>
Reviewer Signature	Date

LABCORE Data Entry Template for Worklist# 26723

Analyst: CRJ Instrument: PU01 20 Book# 28BS7

Method: LA-953-104 Rev/Mod B-0

Worklist Comment: BY-105 FOR @PU23901 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@PU23901	LIQUID		
2	BLNK				@PU23901	LIQUID		
3	SAMPLE	S98T002960 0			@PU23901	LIQUID	98000509	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T								
4	DUP	S98T002960 0			@PU23901	LIQUID		
5	SAMPLE	S98T002961 0			@PU23901	LIQUID	98000509	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T								
6	DUP	S98T002961 0			@PU23901	LIQUID		
7	SAMPLE	S98T002962 0			@PU23901	LIQUID	98000509	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T								
8	DUP	S98T002962 0			@PU23901	LIQUID		
9	SAMPLE	S98T002963 0			@PU23901	LIQUID	98000509	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T								
10	DUP	S98T002963 0			@PU23901	LIQUID		
11	SAMPLE	S98T002964 0			@PU23901	LIQUID	98000509	BY-105 (2)
Analytes Requested: PU23901 , PU23901E, PU23901T								
12	DUP	S98T002964 0			@PU23901	LIQUID		

Final page for worklist # 26723

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID**

			STD
DATE COUNTED	NOV-19-98		0.351
SAMPLE VOLUME in mL	SS	1.000	0.020
SAMPLE DILUTION FACTOR	DF	1.000	0.522
TRACER VOLUME in mL	SPKV	0.100	4157
DIGEST DILUTION FACTOR	DDF	1.000	30
TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
DETECTOR NUMBER	20	Pu 236 cpm	32.810
EFFICIENCY FACTOR	EFF	0.335	1.910
TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	48.870
TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
PU-236 DECAY CORRD VALUE (dpm/mL)	1854.013	Pu 239/240 $\mu\text{Ci/L}$	1.2420E-01
PU-238 TRACER VALUE (dpm/mL)	0.000		
STANDARD BOOK NO	28B57		
STANDARD VALUE in $\mu\text{Ci/mL}$	1.120E-04		
Decay Time = Date Counted - Tracer Preparation Date			
Pu-236 Decay Cor'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time)/1040.95}]			
Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236 * 100 / Pu-236 Decay Cor'd Value * SPKV			
Pu 239/240 $\mu\text{Ci/L}$ = (C239) (Pu 236 Decay Cor'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (2220000 dpm/ μCi)]			
Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)			
Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]			
Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100			
Pu 239/240 $\mu\text{Ci/mL}$	1.24E-04	DETECTION LEVELS in $\mu\text{Ci/mL}$ Pu 239/240 9.11E-06	
Relative Counting Error	= 2.0%		
Pu 236 Tracer Recovery	= 78.3%		

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98

STANDARD.WB1 REV 1.0

943128ML

4821

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

				L	BLNK
Type	DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)		0.865
BLNK	SAMPLE VOLUME in mL	SS 5.000	PU 238 AEA FRAC (C238)		0.000
Work List	SAMPLE DILUTION FACTOR	DF 11.000	PU 239 AEA FRAC (C239)		0.000
26723	TRACER VOLUME in mL	SPKV 0.100	TOTAL AT COUNTS		1680
Test Code	DIGEST DILUTION FACTOR	DDF 1.0000	AT COUNT TIME (MIN)		30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cp (Bkg)		0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm		33.110
LIQUID	EFFICIENCY FACTOR	EFF 0.3347	PU 238 cpm		0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.000
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.00	AEA COUNT TIME		480
Re-run	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.01	Pu 239/240 μ Ci/L =	<	8.173E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.00			

Sample Prep

N/A

Sample # Decay Time = Date Counted - Tracer Preparation Date

WL26723-BLNK Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Instrument Code Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WC16101 Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Prepared By Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

MCB Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) * (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Chemist Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

CRJ

Date Complete

11/19/98	Pu 239/240 μ Ci/mL	< 8.17E-06	DETECTION LEVELS in μ Ci/mL
Analysis Date	Relative Counting Error	= 100.0%	
11/18/98			
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
06:40 PM	Pu 238 μ Ci/mL	< 8.17E-06	8.17E-06
Sample Point	Relative Counting Error	= 100.0%	Pu 238
BY-105(2)	Pu 236 Tracer Recovery	= 78.0%	8.17E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98

BLANK WB1 REV 1.0

943128ML

4922.

WORKBOOK PAGE: SAM3

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

TYPE	DATE COUNTED	NOV-19-98	PU 238 AEA FRAC (C236)	0.870
SAMPLE	SAMPLE VOLUME in mL	SS	5.000 PU 238 AEA FRAC (C238)	0.016
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.046
26723	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1182
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Math	DETECTOR NUMBER	20	PU 238 cpm	22.760
LIQUID	EFFICIENCY FACTOR	EFF	0.335 PU 238 cpm	0.420
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.200
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Rerun	PU-238 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 $\mu\text{Ci/L}$ =	< 1.1490E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Decay Time = Date Counted - Tracer Preparation Date

S98T002960 Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Instrument Code Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WC16101 Pu 239/240 $\mu\text{Ci/L}$ = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Prepared By Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

MCB Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Chemist Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

CRJ

Date Complete

11/19/98

Analysis Date

11/18/98

Analysis Time

06:40 PM

Sample Point

BY-105(2)

Pu 239/240 $\mu\text{Ci/mL}$ < 1.15E-05

Relative Counting Error = 8.4%

DETECTION

LEVELS

in $\mu\text{Ci/mL}$

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 239/240

Pu 238 $\mu\text{Ci/mL}$ < 1.15E-05

1.15E-05

Relative Counting Error = 13.9%

Pu 238

Pu 236 Tracer Recovery = 55.1%

1.15E-05

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 NOV 98

SAMPLE.WB1 REV 1.0

943128ML

4923

WORKBOOK PAGE: DUP4

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)	0.851
SAMPLE VOLUME in mL	5.000	PU 238 AEA FRAC (C238)	0.043
SAMPLE DILUTION FACTOR	11.000	PU 239 AEA FRAC (C239)	0.045
TRACER VOLUME in mL	0.100	TOTAL AT COUNTS	1879
DIGEST DILUTION FACTOR	1.0000	AT COUNT TIME (MIN)	30
TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
DETECTOR NUMBER	20	PU 236 cpm	33.420
EFFICIENCY FACTOR	0.335	PU 238 cpm	1.670
TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.770
TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 $\mu\text{Ci/L}$ =	9.7155E-03
PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002960

Decay Time = Date Counted - Tracer Preparation Date

Instrument Code

WC16101

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Prepared By

MCB

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Chemist

JFR

Pu 239/240 $\mu\text{Ci/L}$ = (C239)(Pu 236 Decay Corr'd Value)(SPKV)/(1000mL)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μCi)]

Analysis

CRJ

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Date Complete

11/19/98

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm)(DF)(DDF)/(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μCi)(D g/L)(SS)]

Analysis Date

11/18/98

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Analysis Time

06:40 PM

Pu 239/240 $\mu\text{Ci/mL}$ = 9.72E-06

Sample Point

BY-105(2)

Relative Counting Error = 6.9%

DETECTION

LEVELS

in $\mu\text{Ci/mL}$

Pu 239/240

7.55E-06

Pu 238

7.55E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: SAM5

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**LIQUID /****SAMPLE**

TYPE	DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)	0.868
SAMPLE	SAMPLE VOLUME in mL SS	5.000	PU 238 AEA FRAC (C238)	0.050
Work List	SAMPLE DILUTION FACTOR DF	11.000	PU 239 AEA FRAC (C239)	0.036
26723	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	1708
Test Code	DIGEST DILUTION FACTOR DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm	31.700
LIQUID	EFFICIENCY FACTOR EFF	0.335	PU 238 cpm	1.810
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.300
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Refun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 μ Ci/L =	< 7.9836E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002961

Instrument Code

WC16101

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Data Complete

11/19/98

Analysis Date

11/18/98

Analysis Time

06:40 PM

Sample Point

BY-105(2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239)(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

11/19/98	Pu 239/240 μ Ci/mL	< 7.98E-06	DETECTION LEVELS
11/18/98	Relative Counting Error	= 8.0%	in μ Ci/mL
06:40 PM	Pu 238 μ Ci/mL	1.06E-05	Pu 239/240
Sample Point	Relative Counting Error	= 6.8%	7.98E-06
BY-105(2)	Pu 236 Tracer Recovery	= 79.5%	Pu 238
			7.98E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

DUP

TYPE	DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)	0.864
DUP	SAMPLE VOLUME in mL SS	5.000	PU 238 AEA FRAC (C238)	0.030
Work Dist	SAMPLE DILUTION FACTOR DF	11.000	PU 239 AEA FRAC (C239)	0.042
26723	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	1694
Test Code	DIGEST DILUTION FACTOR DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm	32.730
LIQUID	EFFICIENCY FACTOR EFF	0.335	PU 238 cpm	1.140
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.580
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Keratin	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 μ Ci/L =	8.9314E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T002961

Instrument Code

WC16101

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date/Complete

11/19/98

Analysis Date

11/18/98

Analysis Time

06:40 PM

Sample Point

BY-105(2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239)(Pu 236 Decay Corr'd Value)(SPKV)(1000mL)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

11/19/98	Pu 239/240 μ Ci/mL	8.93E-06	DETECTION LEVELS
	Relative Counting Error =	7.3%	in μ Ci/mL
11/18/98			Pu 239/240
	NOTE: Pu 238 Result is a LESS THAN Value.		8.12E-06
06:40 PM	Pu 238 μ Ci/mL	< 8.12E-06	Pu 238
	Relative Counting Error =	8.5%	8.12E-06
BY-105(2)	Pu 236 Tracer Recovery =	78.5%	

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: SAM7

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

SAMPLE

TYPE	DATE COUNTED	NOV-19-98	PU 238 AEA FRAC (C236)	0.858
SAMPLE	SAMPLE VOLUME in mL	SS	5.000 PU 238 AEA FRAC (C238)	0.028
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.052
26723	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1759
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 238 cpm	33.460
LIQUID	EFFICIENCY FACTOR	EFF	0.335 PU 238 cpm	1.100
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	2.020
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Result	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 μ Ci/L	1.1135E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep:

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

S98T002962

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WC16101

Pu 239/240 μ Ci/L = (C239)(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μ Ci)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

MCB

Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-238 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]

Chemist

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

CRJ

Data Complete

11/19/98

Pu 239/240 μ Ci/mL 1.11E-05

Analysis Date

Relative Counting Error = 6.5%

11/18/98

Analysis Time

NOTE: Pu 238 Result is a LESS THAN Value.

06:40 PM

Pu 238 μ Ci/mL < 7.93E-06

Sample Point

Relative Counting Error = 8.7%

BY-105(2)

Pu 236 Tracer Recovery = 81.0%

DETECTION

LEVELS

in μ Ci/mL

Pu 239/240

7.93E-06

Pu 238

7.93E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	<i>John Relyea</i>	JFR	Date: 20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

4927

WORKBOOK PAGE: DUP8

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

DUP

Type	DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)	0.838
DUP	SAMPLE VOLUME in mL	SS	5.000 PU 238 AEA FRAC (C238)	0.065
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.052
26723	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1821
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm	32.690
LIQUID	EFFICIENCY FACTOR	EFF	0.335 PU 238 cpm	2.550
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	2.020
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Retun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 µCi/L	1.1401E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		
Sample Prep				
N/A				
Sample #	Decay Time = Date Counted - Tracer Preparation Date			
S98T002962	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]			
Instrument Code	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)			
WC16101	Pu 239/240 µCi/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000mL/L) * (DF) * (DDF) / [(C238) * (SS) * (D g/L) * (2220000 dpm/µCi)]			
Prepared By	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)			
MCB	Pu 238 µCi/L = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/µCi) * (D g/L) * (SS)]			
Chemist	Relative Counting Error = Square Root of {[(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))]} * 1.96 * 100			
JFR				
Analyst				
CRJ				
Date Complete				
11/19/98	Pu 239/240 µCi/mL	1.14E-05	DETECTION LEVELS in µCi/mL	
Analyst Date	Relative Counting Error	= 6.5%		
11/18/98			Pu 239/240	
Analysis Time			8.03E-06	
06:40 PM	Pu 238 µCi/mL	1.43E-05	Pu 238	
Sample Point	Relative Counting Error	= 5.8%	8.03E-06	
BY-105(2)	Pu 236 Tracer Recovery	= 81.9%		

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	JFR	Date:	20 Nov 98
SAMPLE.WB1 REV 1.0	943128ML		

WORKBOOK PAGE: SAM9

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

SAMPLE

TYPE	DATE COUNTED	NOV-19-98	PU 238 AEA FRAC (C238)	0.866
SAMPLE	SAMPLE VOLUME in mL	SS	5.000 PU 238 AEA FRAC (C238)	0.046
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 238 AEA FRAC (C239)	0.027
26723	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1733
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 238 cpm	32.590
LIQUID	EFFICIENCY FACTOR	EFF	0.335 PU 238 cpm	1.720
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.000
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Return	PU-238 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 $\mu\text{Ci/L}$ =	< 7.9047E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

S98T002963

Pu-238 Decay Corr'd Value = Pu-238 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Instrument Code

Pu 238 Tracer Recovery = (Total AT Counts / TC - Bkg) * C238 * 100 / (Pu-238 Decay Corr'd Value * SPKV * EFF)

WC16101

Pu 239/240 $\mu\text{Ci/L}$ = (C239)(Pu 238 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C238)(SS)(D g/L)(2220000 dpm/ μCi)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

MCB

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-238 Tracer Recovery / 100)(2220000 dpm/ μCi)(D g/L)(SS)]

Chemist

Relative Counting Error = Square Root of [(1/(Pu 238 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

CRJ

Data Complete

11/19/98

Pu 239/240 $\mu\text{Ci/mL}$ < 7.90E-06

Analysis Date

Relative Counting Error = 9.1%

11/18/98

Analysis Time

06:40 PM

Pu 238 $\mu\text{Ci/mL}$ 9.76E-06

Sample Points

Relative Counting Error = 7.0%

BY-105(2)

Pu 238 Tracer Recovery = 80.5%

DETECTION

LEVELS

in $\mu\text{Ci/mL}$

Pu 239/240

7.90E-06

Pu 238

7.90E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	<i>John Relyea</i>	Date:	20 Nov 98
	JFR		

SAMPLE.WB1 REV 1.0

943128ML

4929

WORKBOOK PAGE: DUP10

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**LIQUID /****DUP**

Type	DATE COUNTED	NOV-19-98	Pu 238 AEA FRAC (C236)	0.880	
DUP	SAMPLE VOLUME in mL	SS	5.000	Pu 238 AEA FRAC (C238)	0.041
Work List	SAMPLE DILUTION FACTOR	DF	11.000	Pu 239 AEA FRAC (C239)	0.026
26723	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1693
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.070
Matrix	DETECTOR NUMBER	20	Pu 236 cpm		31.800
LIQUID	EFFICIENCY FACTOR	EFF	0.335	Pu 238 cpm	1.480
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm		0.950
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
Reprint	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 $\mu\text{Ci/L}$ =	<	7.8363E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000			
Sample Prep					
N/A					
Sample #	Decay Time = Date Counted - Tracer Preparation Date				
S98T002963	Pu-236 Decay Cor'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]				
Instrument Code	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Cor'd Value * SPKV * EFF)				
WC16101	Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Cor'd Value) * (SPKV / (1000mL/L)) * (DF) * (DDF) / ((C236) * (SS) * (D g/L) * (2220000 dpm/ μCi))				
Prepared By	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)				
MCB	Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-238 Tracer Recovery / 100) * (2220000 dpm/ μCi) * (D g/L) * (SS)]				
Chemist	Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100				
JFR					
Analyst					
CRJ					
Date Complete					
11/19/98	Pu 239/240 $\mu\text{Ci/mL}$	< 7.84E-06	DETECTION LEVELS in $\mu\text{Ci/mL}$		
Analysis Date	Relative Counting Error	= 9.3%			
11/18/98			Pu 239/240		
Analysis Time			7.84E-06		
06:40 PM	Pu 238 $\mu\text{Ci/mL}$	8.56E-06	Pu 238		
Sample Point	Relative Counting Error	= 7.5%	7.84E-06		
BY-105(2)	Pu 236 Tracer Recovery	= 79.9%			

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	<i>John Relyea</i>	JFR	Date: 20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: SAM11

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

SAMPLE

TYPE	DATE COUNTED	NOV-19-98	PU 238 AEA FRAC (C236)	0.859
SAMPLE	SAMPLE VOLUME in mL	SS	5.000 PU 238 AEA FRAC (C238)	0.040
Work List	SAMPLE DILUTION FACTOR	DF	11.000 PU 239 AEA FRAC (C239)	0.038
26723	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1736
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm	31.690
LIQUID	EFFICIENCY FACTOR	EFF	0.335 PU 238 cpm	1.480
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.400
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Return	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	PU 239/240 μ Ci/L =	8.1278E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep:

N/A

Sample #

S98T002964

Instrument Code

WC16101

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Data Complete

11/18/98

Analysis Date

11/18/98

Analysis Time

06:40 PM

Sample Point

BY-105(2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

11/18/98	Pu 239/240 μ Ci/mL	8.13E-06	DETECTION LEVELS
11/18/98	Relative Counting Error =	7.7%	in μ Ci/mL
11/18/98			Pu 239/240
06:40 PM	Pu 238 μ Ci/mL	8.56E-06	8.02E-06
Sample Point	Relative Counting Error =	7.5%	Pu 238
BY-105(2)	Pu 236 Tracer Recovery =	80.0%	8.02E-06

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	<i>John Rolyea</i>	JFR	Date: 26 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

4931

WORKBOOK PAGE: DUP12

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

TYPE	DATE COUNTED	NOV-19-98	PU 236 AEA FRAC (C236)	0.856	
DUP	SAMPLE VOLUME in mL	SS	5.000	PU 238 AEA FRAC (C238)	0.046
Work List	SAMPLE DILUTION FACTOR	DF	11.000	PU 239 AEA FRAC (C239)	0.036
26723	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1751
Digest Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.070
Matrix	DETECTOR NUMBER	20	PU 236 cpm		32.950
LIQUID	EFFICIENCY FACTOR	EFF	0.335	PU 238 cpm	1.780
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		1.370
98005329	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
Refin	PU-236 DECAY CORR'D VALUE (dpm/mL)	1854.013	Pu 239/240 μ Ci/L =	<	8.0072E-03
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample #

S98T002964

Instrument Code

WC16101

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Complete

11/19/98

Analysis Date

11/18/98

Analysis Time

06:40 PM

Sample Point

BY-105(2)

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln 2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 238 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) * (1000 mL/L)] / [(Pu-238 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Relative Counting Error = Square Root of {[(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100}

11/19/98	Pu 239/240 μ Ci/mL	< 8.01E-06	DETECTION LEVELS in μ Ci/mL
	Relative Counting Error =	7.8%	
11/18/98			Pu 239/240
			8.01E-06
06:40 PM	Pu 238 μ Ci/mL	9.87E-06	Pu 238
	Relative Counting Error =	6.9%	8.01E-06
BY-105(2)	Pu 236 Tracer Recovery =	80.4%	

Analyst:	CRJ	Date:	19-Nov-98
Signature of Chemist:	<i>John Relyea</i>	JFR	Date: 20 Nov 98

SAMPLE.WB1 REV 1.0

943128ML

4932

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL26723-STD-PU
File ID: 1a1850.CNF

Counted on: 11/19/98 @ 2:52
Detector: AEA1
Geometry number: 1
Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1152.0	1152.0	361.409	361.409	8.000	2.801	4.000	1.111
2	32.0	32.0	303.561	303.377	16.000	31.428	8.000	11.290
3	69.5	69.5	287.304	287.123	10.000	2.360	5.000	1.690
4	22.6	22.6	270.598	269.452	12.000	5.275	6.000	1.244
5	2046.4	2046.4	229.083	229.082	10.000	2.804	5.000	1.510
6	15.4	15.4	144.505	144.499	136.000	1.000	68.000	0.100
7?	0.0	0.1	134.036	134.036	0.000	0.100	0.000	0.100

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
		Exp.	Obs.	Diff.					
1	Pu236	0.351	5.755	5.754	0.0010.01	32.81	1.6	136.6	0.615E-04
2	Am241	0.020	5.479	5.487	-0.0080.14	1.91	8.7	8.3	0.373E-05
	Pu238		5.487	5.487	0.000			10.8	0.486E-05
3	Th228	0.015	5.400	5.412	-0.0120.01	1.38	9.1	7.9	0.358E-05
4		0.011		5.331	0.02	1.01	11.2	4.1	0.185E-05
5	Pu239	0.522	5.147	5.145	0.0020.01	48.87	1.3	199.3	0.898E-04
	Pu240		5.144	5.145	-0.001			199.3	0.898E-04
6	Np237	0.013	4.769	4.756	0.0130.00	1.22	8.1	5.7	0.257E-05
7		????		4.708		0.00	9236.		
Totals:			0.932 <--valid peaks only-->			87.20			

DETECTOR CALIBRATION

Energy (MEV) = 4.091 + (0.0046) * Channel
Energy range (MeV): 4.091 TO 6.446
Efficiency = 0.2452 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	44908.0	100.000
Smoothed	44908.0	100.000
Composite fit	4933 41858.8	93.210

Residuals

3049.2

6.790

Analyzed by:

SLH2

Spectrum 1a1850.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc   Display Max.: 13090.1

6
6
6
6
6
6
6
7
6
6

.
.
..
...
.5...
.....5..
.....5
.....5
.....5
.....5..

4
4
4
4
4
23
3
2
2
2
2

.1
....1
.....1.
.....1
.....1
.....1
..

Raw Data Dump for AEA Spectrum: 1a1850.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	4.	2.
21	0.	0.	2.	0.	2.	2.	2.	3.	1.	4.
31	1.	0.	1.	2.	3.	3.	0.	0.	5.	2.
41	1.	2.	3.	1.	0.	2.	3.	4.	0.	3.
51	2.	3.	5.	2.	3.	4.	0.	2.	3.	1.
61	1.	2.	4.	0.	4.	2.	2.	6.	3.	5.
71	6.	3.	1.	5.	4.	5.	1.	0.	3.	7.
81	8.	5.	4.	2.	5.	5.	3.	7.	1.	5.
91	8.	1.	3.	2.	3.	2.	4.	0.	8.	9.
101	3.	6.	6.	2.	3.	2.	3.	7.	6.	1.
111	2.	9.	2.	3.	3.	4.	5.	8.	6.	7.
121	12.	4.	12.	8.	10.	10.	5.	8.	6.	9.
131	8.	9.	8.	16.	9.	8.	10.	6.	7.	5.
141	12.	10.	15.	19.	10.	10.	16.	10.	11.	9.
151	17.	18.	13.	18.	14.	16.	18.	14.	11.	16.
161	21.	19.	18.	16.	21.	11.	27.	23.	14.	27.
171	28.	28.	23.	28.	28.	31.	27.	34.	46.	39.
181	41.	41.	51.	50.	39.	39.	56.	60.	63.	47.
191	69.	76.	62.	80.	62.	61.	85.	95.	72.	114.
201	92.	121.	135.	122.	131.	128.	169.	189.	198.	182.
211	214.	227.	280.	306.	328.	433.	481.	546.	643.	697.
221	720.	748.	796.	958.	1157.	1523.	1827.	2243.	2525.	2494.
231	1930.	1177.	420.	109.	19.	7.	3.	7.	5.	8.
241	7.	7.	8.	8.	7.	12.	5.	2.	8.	7.
251	14.	11.	12.	9.	11.	13.	13.	10.	5.	10.
261	18.	10.	7.	17.	15.	26.	30.	30.	24.	30.
271	32.	26.	26.	20.	21.	18.	15.	27.	21.	20.
281	26.	35.	48.	58.	66.	84.	109.	88.	83.	52.
291	34.	16.	20.	23.	18.	28.	25.	24.	25.	32.
301	32.	34.	27.	40.	40.	25.	23.	20.	14.	19.
311	20.	19.	23.	30.	16.	32.	26.	21.	32.	29.
321	37.	39.	38.	37.	59.	57.	51.	40.	45.	59.
331	57.	64.	57.	62.	82.	88.	89.	104.	91.	116.
341	137.	116.	172.	204.	199.	204.	243.	315.	364.	429.
351	544.	641.	624.	579.	512.	621.	613.	771.	865.	1168.
361	1408.	1437.	1265.	793.	319.	91.	21.	1.	4.	0.
371	0.	1.	3.	0.	1.	0.	0.	0.	0.	2.
381	0.	0.	0.	2.	0.	2.	0.	1.	0.	2.
391	0.	0.	1.	1.	0.	0.	0.	0.	0.	0.
401	2.	2.	0.	0.	0.	0.	0.	1.	0.	0.
411	0.	1.	1.	0.	0.	0.	0.	1.	0.	0.
421	0.	0.	1.	1.	0.	0.	0.	1.	2.	0.
431	1.	2.	1.	0.	1.	0.	0.	1.	0.	2.
441	0.	0.	1.	0.	0.	0.	0.	1.	1.	0.
451	1.	0.	0.	0.	1.	0.	1.	0.	0.	3.
461	2.	1.	0.	1.	1.	1.	0.	0.	2.	3.
471	3.	7.	7.	12.	5.	5.	3.	2.	3.	1.
481	2.	1.	0.	0.	0.	0.	0.	1.	0.	1.
491	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL26723-BLNK-PU
 File ID: 2a2047.CNF

Counted on: 11/19/98 @ 2:53
 Detector: AEA2
 Geometry number: 1
 Count time: 28810. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1077.7	1077.7	362.264	362.264	12.000	3.107	6.000	1.127
2	70.8	70.8	287.636	287.616	12.000	7.268	6.000	2.417

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate c/m	%err @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.				
1	Pu236	0.865	5.755	5.753	0.0020.01	33.11	1.6	145.4	0.655E-04
2	Th228	0.071	5.400	5.409	-0.0090.03	2.73	5.5	16.5	0.745E-05
Totals:		0.936	<--valid peaks only-->			35.83			

DETECTOR CALIBRATION

Energy (MeV) = $4.086 + (0.0046) * \text{Channel}$
 Energy range (MeV): 4.086 TO 6.441
 Efficiency = 0.2324 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18379.0	100.000
Smoothed	18379.0	100.000
Composite fit	17206.7	93.621
Residuals	1172.3	6.379

Analyzed by:

SLH2

```

2.
.2
....2
...2
..
.
1.
.1..
....1.
.....1
.....1.
.....1
.....1
.....1
.....1
.....1

```

Raw Data Dump for AEA Spectrum: 2a2047.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	0.	0.	0.	0.	0.	0.	2.	0.	0.	1.
31	0.	1.	1.	0.	0.	1.	1.	0.	1.	0.
41	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	1.	0.	0.	0.	1.
61	0.	1.	0.	1.	0.	1.	1.	0.	0.	0.
71	0.	0.	0.	1.	0.	0.	0.	1.	0.	1.
81	0.	1.	0.	0.	0.	0.	0.	1.	2.	0.
91	1.	1.	0.	1.	0.	1.	0.	0.	1.	1.
101	0.	1.	1.	0.	0.	2.	1.	0.	0.	0.
111	1.	1.	0.	0.	0.	2.	0.	0.	0.	0.
121	1.	0.	1.	0.	1.	2.	1.	2.	0.	1.
131	0.	0.	0.	0.	0.	1.	2.	3.	0.	1.
141	2.	0.	0.	0.	2.	0.	0.	1.	0.	2.
151	0.	1.	2.	1.	1.	2.	1.	0.	1.	0.
161	1.	3.	5.	2.	0.	1.	2.	1.	2.	0.
171	1.	3.	0.	1.	1.	1.	1.	2.	2.	2.
181	1.	3.	1.	1.	0.	2.	2.	0.	1.	2.
191	3.	4.	1.	0.	2.	3.	4.	1.	1.	5.
201	3.	1.	2.	2.	3.	0.	1.	5.	3.	3.
211	2.	3.	4.	6.	2.	2.	2.	4.	4.	8.
221	4.	8.	3.	5.	6.	5.	4.	3.	4.	5.
231	12.	5.	0.	7.	4.	2.	3.	5.	6.	4.
241	3.	4.	7.	9.	4.	6.	6.	7.	8.	10.
251	4.	10.	5.	10.	6.	8.	9.	14.	18.	5.
261	12.	17.	15.	18.	15.	24.	18.	19.	23.	29.
271	27.	38.	24.	17.	23.	19.	21.	28.	20.	32.
281	30.	31.	36.	46.	80.	76.	78.	84.	76.	64.
291	43.	36.	12.	21.	17.	15.	20.	25.	23.	22.
301	24.	24.	20.	20.	28.	25.	19.	27.	17.	37.
311	28.	25.	19.	19.	26.	38.	28.	29.	33.	44.
321	42.	40.	45.	42.	50.	42.	58.	51.	67.	68.
331	63.	62.	88.	85.	70.	81.	94.	104.	90.	134.
341	108.	175.	166.	157.	206.	224.	225.	302.	314.	402.
351	470.	567.	604.	626.	587.	535.	588.	673.	761.	881.
361	1122.	1273.	1355.	1099.	702.	339.	138.	35.	4.	1.
371	0.	0.	1.	0.	0.	2.	1.	0.	0.	1.
381	0.	0.	0.	1.	1.	1.	0.	0.	1.	0.
391	1.	1.	0.	0.	1.	0.	0.	1.	0.	0.
401	0.	0.	2.	0.	0.	1.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	1.	1.	1.	1.	0.	0.	1.	1.	1.	0.
431	2.	1.	0.	1.	2.	0.	0.	0.	0.	1.
441	1.	0.	1.	0.	1.	0.	0.	1.	2.	0.
451	1.	0.	0.	0.	0.	0.	0.	1.	2.	1.
461	2.	0.	2.	2.	0.	1.	2.	4.	3.	4.
471	10.	2.	6.	4.	3.	2.	7.	8.	3.	1.
481	4.	1.	0.	0.	0.	0.	1.	0.	0.	0.
491	2.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98R2960-SAM-PU
File ID: 3a3338.CNF

Counted on: 11/19/98 @ 2:53
Detector: AEA3
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	829.3	829.3	361.352	361.352	10.000	3.164	5.000	1.390
2	23.1	23.1	301.956	301.771	10.000	0.714	5.000	0.398
3	44.3	44.3	286.663	286.551	10.000	3.139	5.000	2.142
4	13.5	13.5	269.460	269.028	10.000	3.301	5.000	1.429
5	43.0	43.0	228.060	228.055	10.000	4.563	5.000	2.311

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Exp.	Centroid Obs.	Diff.	FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
1	Pu236	0.870	5.755	5.754	0.0010	0.01	22.76	1.9	97.1	0.437E-04
2	Pu238	0.016	5.487	5.480	0.0070	0.00	0.42	14.2	2.4	0.109E-05
	Am241		5.479	5.480	-0.001				1.9	0.838E-06
3	Th228	0.037	5.400	5.410	-0.0100	0.01	0.98	9.8	5.7	0.259E-05
4		0.014		5.329		0.02	0.38	18.1	1.6	0.709E-06
5	Pu239	0.046	5.147	5.140	0.0070	0.02	1.20	8.2	5.0	0.225E-05
	Pu240		5.144	5.140	0.004				5.0	0.225E-05

Totals: 0.984 <--valid peaks only--> 25.73

DETECTOR CALIBRATION

Energy (MeV) = 4.091 + (0.0046)*Channel
Energy range (MeV): 4.091 TO 6.447
Efficiency = 0.2393 CPM/DCM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	12556.0	100.000
Smoothed	12556.0	100.000
Composite fit	12352.6	98.380
Residuals	203.4	1.620

4940

Analyzed by: _____

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 4091.9

5
.5
....5
5
4
4
4
4
...3
3.
.2

.
1
..1
..... 1
.....1..
..... 1
.....1
..1...

Raw Data Dump for AEA Spectrum: 3a3338.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	1.	0.	0.	0.	0.	1.	0.	0.
31	0.	0.	0.	0.	1.	0.	1.	0.	1.
41	0.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	1.	0.
61	0.	0.	0.	0.	0.	1.	0.	0.	1.
71	2.	0.	0.	0.	0.	0.	0.	1.	0.
81	0.	1.	0.	1.	0.	0.	0.	0.	0.
91	0.	1.	1.	0.	0.	0.	0.	0.	1.
101	0.	0.	1.	1.	0.	0.	0.	0.	0.
111	1.	0.	0.	0.	0.	0.	1.	0.	0.
121	0.	0.	1.	2.	0.	0.	0.	0.	1.
131	0.	0.	0.	0.	0.	0.	0.	1.	2.
141	2.	1.	1.	1.	0.	0.	0.	1.	3.
151	0.	0.	2.	1.	1.	0.	2.	0.	0.
161	1.	0.	0.	2.	0.	2.	1.	0.	0.
171	3.	0.	1.	0.	0.	0.	3.	0.	0.
181	0.	0.	0.	0.	2.	0.	1.	2.	0.
191	0.	0.	0.	2.	3.	2.	2.	1.	1.
201	2.	3.	2.	1.	3.	5.	2.	5.	3.
211	6.	5.	1.	6.	14.	13.	12.	14.	22.
221	8.	19.	29.	30.	32.	37.	58.	40.	57.
231	24.	28.	10.	8.	6.	0.	2.	2.	2.
241	3.	3.	0.	3.	2.	1.	4.	2.	4.
251	1.	3.	3.	2.	1.	2.	6.	4.	2.
261	7.	5.	11.	7.	14.	17.	12.	15.	22.
271	22.	10.	10.	7.	9.	8.	12.	10.	9.
281	25.	28.	32.	47.	42.	51.	56.	60.	29.
291	19.	5.	10.	8.	8.	5.	10.	8.	8.
301	10.	13.	13.	10.	6.	10.	8.	5.	4.
311	6.	5.	7.	3.	8.	12.	9.	10.	11.
321	7.	10.	12.	13.	6.	18.	12.	13.	19.
331	27.	18.	20.	32.	29.	29.	42.	31.	49.
341	62.	67.	70.	84.	105.	134.	157.	204.	254.
351	406.	381.	417.	391.	390.	426.	428.	600.	715.
361	980.	1026.	864.	615.	296.	108.	18.	4.	3.
371	0.	0.	0.	0.	1.	0.	2.	0.	1.
381	0.	0.	2.	0.	0.	0.	0.	1.	1.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	1.	0.	0.	0.	0.	1.
421	0.	0.	3.	0.	1.	1.	1.	1.	1.
431	2.	0.	2.	0.	0.	0.	3.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	1.	0.	0.	0.	0.	0.	0.
461	0.	2.	0.	1.	0.	1.	3.	1.	4.
471	6.	0.	2.	12.	6.	13.	5.	6.	6.
481	2.	1.	0.	0.	0.	0.	1.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

HNF-1663 REV. 0

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2960-DUP-PU
File ID: 4a4339.CNF

Counted on: 11/19/98 @ 2:54
Detector: AEA4
Geometry number: 1
Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1218.0	1218.0	360.286	360.286	10.000	2.774	5.000	1.166
2	20.8	20.8	301.876	301.695	80.000	1.000	40.000	0.100
3	43.8	43.8	285.706	285.607	10.000	2.077	5.000	1.966
4?	12.0	12.0	268.289	267.605	16.000	3.382	8.000	9.374
5	72.5	72.5	227.484	227.471	12.000	3.656	6.000	2.159

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.851	5.755	5.744	0.0110.01	33.42	1.5	148.3	0.668E-04
2	Pu238	0.043	5.487	5.475	0.0120.00	1.67	7.0	10.1	0.455E-05
	Am241		5.479	5.475	0.004			7.7	0.349E-05
3	Th228	0.019	5.400	5.401	-0.0010.01	0.75	17.1	4.6	0.206E-05
4	????			5.318		0.22	44.0		
5	Pu239	0.045	5.147	5.133	0.0140.02	1.77	7.2	7.7	0.347E-05
	Pu240		5.144	5.133	0.011			7.7	0.347E-05
Totals:			0.958	<--valid peaks only-->		37.61			

DETECTOR CALIBRATION

Energy (MEV) = 4.087 + (0.0046)*Channel
Energy range (MeV): 4.087 TO 6.442
Efficiency = 0.2299 CPM/DEPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18858.0	100.000
Smoothed	18857.9	100.000
Composite fit	18162.1	96.310
Residuals	695.9	3.690

Spectrum 4a4339.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,.., etc

Display Max.:

6508.0

```

.
.5
2...5
2
2
2
4
4
3
23..
2
2
.
..
1.
...1
..... 1
.....1....
.....1
.....1.
...

```

Raw Data Dump for AEA Spectrum: 4a4339.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	0.	0.	0.	1.	2.	0.	1.	1.	0.	0.
31	0.	0.	1.	0.	1.	0.	0.	1.	0.	0.
41	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	0.	1.	0.	1.
61	0.	0.	0.	0.	1.	1.	0.	0.	0.	2.
71	1.	1.	0.	0.	1.	1.	0.	0.	0.	0.
81	0.	0.	0.	1.	1.	1.	0.	1.	0.	2.
91	0.	1.	0.	0.	0.	0.	1.	1.	1.	1.
101	0.	1.	2.	0.	1.	1.	2.	0.	1.	1.
111	0.	0.	0.	2.	1.	0.	3.	1.	1.	2.
121	0.	1.	1.	0.	1.	1.	0.	0.	1.	0.
131	0.	0.	0.	2.	0.	0.	4.	2.	3.	2.
141	2.	1.	3.	2.	1.	2.	4.	2.	3.	3.
151	1.	2.	0.	2.	0.	1.	0.	2.	1.	3.
161	2.	0.	1.	0.	0.	1.	2.	2.	2.	0.
171	0.	1.	4.	1.	4.	2.	1.	1.	2.	5.
181	1.	3.	3.	1.	3.	1.	2.	3.	2.	3.
191	1.	3.	2.	1.	2.	3.	2.	3.	3.	2.
201	1.	7.	2.	5.	6.	3.	8.	8.	11.	10.
211	15.	11.	21.	15.	15.	13.	23.	18.	24.	24.
221	31.	37.	35.	56.	54.	82.	95.	94.	72.	59.
231	34.	35.	10.	8.	2.	2.	6.	2.	5.	5.
241	5.	4.	3.	5.	4.	2.	5.	0.	6.	6.
251	4.	7.	6.	7.	5.	5.	8.	5.	5.	12.
261	8.	12.	10.	10.	20.	28.	24.	21.	21.	25.
271	16.	17.	16.	11.	10.	16.	17.	18.	23.	22.
281	29.	35.	51.	57.	73.	67.	63.	56.	23.	12.
291	8.	9.	16.	16.	22.	16.	16.	16.	8.	26.
301	29.	20.	21.	25.	18.	15.	12.	13.	13.	17.
311	19.	22.	19.	26.	31.	25.	20.	29.	24.	21.
321	28.	28.	21.	37.	41.	36.	56.	45.	48.	58.
331	50.	42.	52.	61.	62.	78.	81.	97.	104.	112.
341	121.	117.	147.	208.	201.	240.	335.	393.	443.	589.
351	639.	646.	638.	557.	575.	662.	757.	1046.	1240.	1475.
361	1558.	1240.	799.	293.	61.	10.	2.	1.	1.	3.
371	4.	0.	1.	0.	0.	0.	0.	0.	1.	2.
381	0.	0.	0.	0.	0.	1.	0.	2.	0.	1.
391	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	1.	0.	0.	0.	1.
411	0.	1.	0.	1.	0.	0.	0.	0.	0.	1.
421	1.	0.	1.	0.	1.	0.	1.	0.	1.	2.
431	0.	1.	1.	0.	0.	0.	0.	0.	0.	0.
441	2.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	2.	0.	0.	0.	1.	0.	0.	0.	0.	0.
461	0.	0.	0.	0.	1.	4.	2.	2.	8.	3.
471	4.	5.	8.	2.	6.	5.	3.	2.	1.	1.
481	2.	0.	1.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	1.	0.	1.	1.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2961-SAM-PU
File ID: 5a5272.CNF

Counted on: 11/19/98 @ 2:55
Detector: AEA5
Geometry number: 1
Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1150.7	1150.7	363.367	363.367	10.000	2.769	5.000	1.156
2	22.9	22.9	303.916	303.677	42.000	62.493	21.000	27.564
3	40.3	40.3	288.457	288.233	10.000	1.712	5.000	0.739
4	51.9	51.9	230.052	230.050	10.000	3.699	5.000	2.094

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA	Peak Centroid			Count Rate	%err	d/m	Activity
		Frac	Exp.	Obs.	Diff.	c/m	@95		uCi/ea
1	Pu236	0.868	5.755	5.762	-.0070.01	31.70	1.6	133.4	0.601E-04
	Cm243		5.779	5.762	0.017			179.1	0.807E-04
2	Am241	0.050	5.479	5.488	-.0090.29	1.81	12.9	8.0	0.358E-05
	Pu238		5.487	5.488	-.001			10.4	0.468E-05
3	Th228	0.027	5.400	5.417	-.0170.01	1.00	11.3	5.8	0.261E-05
4	Pu239	0.036	5.147	5.149	-.0020.02	1.30	7.9	5.4	0.241E-05
	Pu240		5.144	5.149	-.005			5.4	0.241E-05

Totals:		0.981	<---valid peaks only-->			35.80			

DETECTOR CALIBRATION

Energy (MEV) = 4.091 + (0.0046)*Channel

Energy range (MeV): 4.091 TO 6.446

Efficiency = 0.2424 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17524.0	100.000
Smoothed	17524.0	100.000
Composite fit	17186.9	98.076
Residuals	337.1	1.924

Analyzed by:

SLH2

Spectrum 5a5272.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 6963.8

4
 ..4
 4

°
 3
 23
 3.
 2
 2
 2
 2.
 21.

.....1
1
1
1
 ....1....

Raw Data Dump for AEA Spectrum: 5a5272.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
31	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
41	1.	1.	0.	0.	1.	0.	0.	0.	0.	0.
51	0.	1.	0.	0.	0.	1.	0.	0.	0.	0.
61	0.	0.	0.	1.	0.	0.	1.	0.	0.	0.
71	0.	0.	0.	1.	0.	0.	0.	0.	0.	2.
81	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
91	0.	0.	2.	0.	0.	0.	0.	0.	0.	0.
101	0.	1.	0.	0.	1.	1.	0.	2.	1.	0.
111	1.	0.	0.	1.	1.	0.	2.	0.	0.	1.
121	0.	1.	1.	1.	2.	1.	2.	1.	1.	1.
131	0.	2.	0.	0.	0.	1.	2.	3.	0.	0.
141	2.	3.	0.	0.	3.	1.	0.	0.	1.	2.
151	0.	1.	1.	1.	1.	6.	1.	1.	0.	3.
161	0.	0.	1.	0.	1.	1.	1.	1.	0.	1.
171	0.	2.	0.	1.	2.	1.	1.	0.	3.	2.
181	0.	1.	3.	2.	1.	1.	1.	3.	1.	2.
191	0.	1.	2.	2.	4.	2.	6.	5.	6.	6.
201	4.	3.	4.	2.	6.	8.	2.	5.	8.	6.
211	7.	7.	7.	13.	5.	11.	10.	6.	9.	11.
221	17.	18.	20.	17.	28.	38.	34.	60.	50.	58.
231	71.	45.	34.	15.	11.	4.	3.	4.	3.	1.
241	2.	0.	4.	1.	5.	6.	5.	2.	2.	0.
251	2.	7.	5.	7.	2.	6.	6.	3.	6.	5.
261	4.	8.	8.	10.	11.	12.	10.	20.	18.	17.
271	12.	21.	14.	18.	12.	20.	20.	14.	8.	14.
281	19.	24.	24.	32.	35.	39.	61.	69.	63.	55.
291	44.	24.	7.	10.	13.	11.	17.	13.	18.	14.
301	22.	16.	21.	28.	21.	16.	18.	19.	16.	12.
311	15.	15.	15.	10.	10.	21.	18.	18.	19.	27.
321	11.	17.	22.	40.	27.	35.	32.	36.	29.	38.
331	45.	42.	56.	43.	47.	53.	52.	70.	62.	88.
341	70.	69.	110.	115.	130.	134.	187.	202.	271.	286.
351	344.	463.	566.	631.	612.	555.	514.	529.	613.	692.
361	941.	1198.	1420.	1401.	1302.	742.	302.	75.	9.	1.
371	0.	1.	1.	1.	0.	0.	0.	1.	0.	0.
381	1.	0.	2.	0.	1.	1.	0.	1.	0.	0.
391	0.	0.	0.	1.	2.	0.	0.	0.	0.	0.
401	0.	0.	1.	1.	0.	0.	0.	0.	0.	1.
411	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
421	0.	2.	1.	0.	0.	1.	0.	1.	1.	2.
431	3.	0.	0.	1.	1.	0.	0.	0.	0.	0.
441	0.	0.	0.	1.	0.	0.	1.	0.	1.	0.
451	1.	0.	0.	0.	0.	1.	1.	0.	1.	1.
461	0.	0.	0.	0.	0.	1.	1.	1.	1.	2.
471	3.	0.	3.	7.	3.	6.	3.	3.	11.	5.
481	3.	3.	1.	1.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2961-DUP-PU
File ID: 6a6277.CNF

Counted on: 11/19/98 @ 2:55
Detector: AEA6
Geometry number: 1
Count time: 28800. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1386.8	1386.8	360.350	360.350	8.000	2.476	4.000	1.292
2	21.0	21.0	302.434	302.392	14.000	17.897	7.000	4.688
3	54.7	54.7	285.728	285.645	10.000	2.201	5.000	2.650
4	16.9	16.9	268.676	268.496	8.000	3.617	4.000	1.878
5	70.2	70.2	227.637	227.634	10.000	2.920	5.000	1.812

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.864	5.755	5.751	0.0040.01	32.73	1.6	144.6	0.651E-04
2	Am241	0.030	5.479	5.484	-0.0050.08	1.14	9.0	5.3	0.237E-05
	Pu238		5.487	5.484	0.003			6.9	0.310E-05
3	Th228	0.023	5.400	5.407	-0.0070.01	0.89	11.6	5.4	0.244E-05
4		0.012		5.328	0.02	0.44	16.9	1.9	0.851E-06
5	Pu239	0.042	5.147	5.140	0.0070.01	1.58	7.1	6.8	0.308E-05
	Pu240		5.144	5.140	0.004			6.8	0.308E-05

Totals:			0.970	<--valid peaks only-->		36.78			

DETECTOR CALIBRATION

Energy (MeV) = 4.093 + (0.0046)*Channel
Energy range (MeV): 4.093 TO 6.448
Efficiency = 0.2310 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18194.0	100.000
Smoothed	18193.8	99.999
Composite fit	17654.9	97.037
Residuals	539.1	2.963

4949 Analyzed by: _____

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

6740.6

.5
...5

4
4
4
2.3
3
2
2
2
.

.1
.....1

.....1....

.....1..

...

1

Raw Data Dump for AEA Spectrum: 6a6277.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	1.	0.	1.	0.	0.	0.	1.	0.	0.	0.
31	0.	0.	1.	0.	0.	1.	0.	0.	0.	0.
41	0.	0.	0.	0.	1.	1.	0.	0.	0.	3.
51	0.	0.	0.	0.	1.	0.	0.	1.	0.	0.
61	0.	0.	1.	0.	0.	2.	0.	0.	0.	1.
71	0.	0.	0.	1.	0.	0.	2.	0.	0.	2.
81	0.	0.	1.	1.	0.	0.	0.	0.	0.	0.
91	0.	0.	1.	0.	0.	1.	2.	0.	1.	3.
101	1.	1.	0.	0.	2.	0.	0.	0.	0.	0.
111	1.	1.	0.	1.	1.	1.	2.	0.	2.	0.
121	0.	1.	1.	0.	1.	1.	1.	0.	0.	2.
131	1.	2.	1.	0.	0.	1.	0.	1.	2.	0.
141	1.	2.	1.	3.	0.	1.	3.	0.	5.	2.
151	1.	2.	2.	3.	2.	1.	1.	0.	0.	2.
161	2.	0.	2.	3.	0.	1.	0.	3.	1.	3.
171	2.	2.	3.	2.	2.	3.	2.	0.	1.	1.
181	4.	3.	0.	3.	4.	1.	1.	1.	1.	1.
191	3.	3.	6.	0.	1.	2.	3.	2.	4.	1.
201	3.	1.	2.	4.	5.	4.	2.	5.	2.	6.
211	7.	4.	9.	10.	10.	17.	20.	16.	18.	19.
221	26.	36.	25.	61.	63.	52.	84.	94.	74.	53.
231	26.	8.	8.	2.	1.	3.	3.	3.	1.	5.
241	6.	2.	1.	1.	2.	1.	1.	6.	0.	6.
251	4.	4.	3.	5.	0.	6.	2.	9.	7.	4.
261	13.	8.	5.	18.	11.	18.	21.	22.	18.	31.
271	11.	10.	10.	5.	12.	6.	6.	16.	16.	15.
281	20.	30.	39.	66.	86.	72.	83.	44.	23.	11.
291	9.	13.	13.	10.	9.	10.	7.	10.	14.	23.
301	17.	30.	18.	27.	13.	11.	12.	8.	8.	10.
311	9.	10.	11.	12.	12.	13.	19.	15.	9.	11.
321	22.	12.	16.	15.	17.	22.	21.	29.	32.	33.
331	39.	40.	32.	33.	33.	38.	45.	56.	47.	71.
341	88.	90.	100.	111.	163.	170.	245.	328.	420.	575.
351	681.	654.	543.	477.	519.	657.	780.	1071.	1476.	1691.
361	1846.	1464.	820.	254.	46.	4.	1.	0.	0.	1.
371	0.	0.	0.	0.	0.	0.	1.	0.	1.	1.
381	1.	0.	0.	0.	0.	0.	0.	0.	1.	0.
391	1.	0.	1.	0.	0.	0.	0.	0.	0.	1.
401	0.	0.	2.	0.	0.	0.	0.	0.	1.	1.
411	2.	1.	0.	0.	0.	0.	0.	1.	0.	0.
421	1.	1.	0.	1.	0.	0.	0.	0.	0.	1.
431	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	1.	1.	1.	0.
451	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
461	2.	0.	2.	0.	1.	1.	1.	5.	1.	4.
471	1.	5.	4.	5.	5.	9.	9.	1.	3.	5.
481	3.	0.	0.	0.	1.	0.	0.	0.	0.	0.
491	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2962-SAM-PU
File ID: 7a7303.CNF

Counted on: 11/19/98 @ 2:56
Detector: AEA7
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1238.5	1238.5	362.298	362.298	8.000	2.492	4.000	1.039
2	22.0	22.0	303.968	303.760	14.000	16.327	7.000	4.886
3	54.0	54.0	287.402	287.286	10.000	2.268	5.000	1.767
4	15.5	15.5	270.233	269.714	10.000	3.662	5.000	1.219
5	86.3	86.3	229.077	229.071	10.000	3.076	5.000	1.810

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		FWHM	Count Rate	%err c/m	d/m	Activity uCi/ea	
			Exp.	Obs. Diff.						
1	Pu236	0.858	5.755	5.762	-0.0070	0.1	33.46	1.5	139.0	0.626E-04
	Cm243		5.779	5.762	0.017				186.6	0.841E-04
2	Am241	0.028	5.479	5.492	-0.0130	0.8	1.10	10.2	4.8	0.216E-05
	Pu238		5.487	5.492	-0.005				6.2	0.281E-05
3	Th228	0.026	5.400	5.417	-0.0170	0.1	1.02	10.4	5.9	0.264E-05
4		0.013		5.336	0.02		0.52	15.5	2.1	0.950E-06
5	Pu239	0.052	5.147	5.149	-0.0020	0.1	2.02	6.3	8.2	0.370E-05
	Pu240		5.144	5.149	-0.005				8.2	0.370E-05
Totals:			0.978	<--valid peaks only-->			38.13			

DETECTOR CALIBRATION

Energy (MeV) = 4.095 + (0.0046)*Channel

Energy range (MeV): 4.095 TO 6.450

Efficiency = 0.2456 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18722.0	100.000
Smoothed	18721.6	99.998
Composite fit	18301.9	97.756
Residuals	420.1	2.244

Spectrum 7a7303.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

6558.9

```

.5
.....5
5

4
4
4
4
2.3
3.
2
2
2
1.
..1.
.....1
.....1...
.....1
.....1
1....

```

Raw Data Dump for AEA Spectrum: 7a7303.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	1.	0.	1.	0.	0.	0.	2.	0.	1.	0.
31	0.	0.	0.	0.	1.	1.	0.	0.	0.	0.
41	0.	1.	1.	0.	0.	0.	0.	0.	1.	0.
51	0.	1.	0.	0.	0.	1.	0.	2.	0.	2.
61	2.	2.	0.	0.	1.	0.	1.	0.	0.	2.
71	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
81	0.	2.	0.	1.	0.	0.	0.	1.	0.	1.
91	0.	0.	0.	3.	0.	1.	0.	2.	0.	0.
101	0.	1.	0.	1.	1.	0.	1.	1.	1.	1.
111	0.	1.	0.	0.	1.	1.	0.	1.	1.	1.
121	1.	1.	0.	1.	0.	1.	0.	0.	0.	2.
131	0.	0.	1.	0.	2.	2.	1.	2.	1.	2.
141	0.	0.	1.	0.	1.	1.	3.	0.	1.	4.
151	1.	2.	0.	1.	0.	3.	2.	0.	1.	1.
161	0.	1.	2.	1.	1.	1.	0.	5.	2.	2.
171	4.	3.	2.	0.	1.	1.	0.	1.	2.	2.
181	2.	1.	3.	1.	1.	2.	5.	3.	0.	5.
191	5.	1.	3.	3.	3.	2.	5.	3.	1.	5.
201	12.	6.	8.	6.	5.	5.	4.	13.	7.	8.
211	5.	8.	18.	9.	11.	16.	18.	21.	25.	25.
221	32.	25.	39.	36.	50.	58.	79.	77.	119.	104.
231	68.	51.	22.	17.	12.	3.	0.	1.	2.	7.
241	1.	4.	3.	6.	2.	5.	1.	2.	6.	3.
251	5.	5.	6.	6.	4.	5.	8.	6.	9.	8.
261	11.	5.	10.	9.	13.	8.	14.	18.	17.	26.
271	19.	16.	15.	9.	18.	7.	13.	12.	15.	14.
281	28.	15.	32.	45.	47.	60.	87.	62.	67.	40.
291	25.	6.	10.	14.	11.	9.	17.	7.	15.	13.
301	19.	23.	26.	33.	17.	23.	21.	17.	15.	7.
311	12.	10.	16.	13.	13.	20.	21.	13.	25.	18.
321	25.	22.	23.	27.	26.	41.	25.	25.	30.	48.
331	48.	52.	61.	46.	62.	57.	62.	86.	71.	97.
341	120.	124.	125.	139.	160.	201.	201.	251.	303.	446.
351	514.	629.	679.	637.	569.	530.	518.	603.	769.	983.
361	1263.	1484.	1650.	1324.	647.	219.	46.	8.	3.	0.
371	2.	0.	0.	0.	0.	1.	1.	1.	0.	0.
381	0.	0.	0.	0.	1.	0.	0.	2.	0.	0.
391	0.	1.	0.	0.	1.	0.	1.	0.	1.	0.
401	0.	0.	0.	1.	0.	2.	0.	0.	1.	0.
411	0.	0.	0.	0.	1.	0.	0.	0.	0.	2.
421	1.	1.	1.	0.	0.	0.	1.	0.	0.	1.
431	0.	0.	1.	0.	0.	1.	1.	1.	1.	0.
441	1.	0.	0.	0.	0.	0.	0.	1.	1.	0.
451	1.	1.	0.	1.	1.	0.	2.	0.	0.	1.
461	1.	0.	1.	1.	1.	1.	4.	0.	1.	0.
471	2.	2.	3.	7.	3.	7.	3.	4.	5.	3.
481	3.	5.	3.	1.	0.	0.	0.	0.	0.	0.
491	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2962-DUP-PU
File ID: 8a8283.CNF

Counted on: 11/19/98 @ 2:56
Detector: AEA8
Geometry number: 1
Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1269.5	1269.5	362.801	362.801	10.000	2.437	5.000	1.090
2	25.5	25.5	303.443	303.342	38.000	100.189	19.000	35.606
3	49.6	49.6	288.081	287.944	8.000	1.593	4.000	1.686
4	18.9	18.9	269.811	269.501	8.000	2.858	4.000	1.754
5	80.5	80.5	229.427	229.425	10.000	3.508	5.000	1.909

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err	Activity
			Exp.	Obs. Diff.	c/m	@95	d/m uCi/ea
1	Pu236	0.838	5.755	5.762 -0.0070.01	32.69	1.6	147.2 0.663E-04
	Cm243		5.779	5.762 0.017			197.6 0.890E-04
2	Am241	0.065	5.479	5.489 -0.0100.46	2.55	17.4	12.0 0.540E-05
	Pu238		5.487	5.489 -0.002			15.6 0.705E-05
3	Th228	0.019	5.400	5.418 -0.0180.01	0.75	15.5	4.6 0.209E-05
4		0.011		5.333 0.01	0.43	18.6	1.9 0.846E-06
5	Pu239	0.052	5.147	5.149 -0.0020.02	2.02	6.3	8.9 0.402E-05
	Pu240		5.144	5.149 -0.005			8.9 0.402E-05

Totals: 0.985 <--valid peaks only--> 38.44

DETECTOR CALIBRATION

Energy (MEV) = 4.093 + (0.0046)*Channel

Energy range (MeV): 4.093 TO 6.448

Efficiency = 0.2267 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18728.0	100.000
Smoothed	18727.9	100.000
Composite fit	18455.0	98.542
Residuals	273.0	1.458

Spectrum 8a8283.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 7023.4

```

.5
....5
5

4
4
4
23.
3.
2
2
2
2.
21.
2.....1
.....1..
..... 1
.....1
.1....

```

Raw Data Dump for AEA Spectrum: 8a8283.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	1.	1.	0.	0.	1.	1.	1.	0.	0.	0.
31	0.	0.	0.	1.	0.	1.	0.	1.	0.	0.
41	0.	0.	0.	0.	1.	0.	0.	1.	1.	0.
51	1.	0.	1.	0.	1.	0.	0.	0.	1.	1.
61	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
71	0.	0.	0.	1.	0.	0.	0.	1.	1.	0.
81	0.	1.	1.	0.	0.	1.	0.	0.	2.	0.
91	0.	0.	0.	0.	1.	0.	0.	0.	1.	4.
101	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
111	0.	0.	0.	2.	1.	1.	0.	0.	0.	0.
121	0.	0.	0.	0.	1.	1.	3.	1.	0.	0.
131	1.	1.	3.	1.	0.	0.	2.	1.	5.	2.
141	1.	2.	3.	1.	1.	1.	3.	1.	0.	3.
151	2.	1.	0.	2.	1.	1.	1.	1.	0.	0.
161	0.	2.	0.	0.	0.	2.	0.	0.	2.	2.
171	1.	2.	2.	3.	0.	1.	1.	2.	1.	1.
181	2.	2.	2.	0.	4.	5.	2.	2.	3.	3.
191	4.	0.	0.	3.	2.	4.	2.	1.	7.	4.
201	2.	3.	5.	4.	3.	5.	6.	8.	8.	7.
211	8.	4.	7.	15.	15.	13.	16.	22.	21.	32.
221	24.	30.	45.	31.	44.	52.	63.	100.	97.	90.
231	99.	56.	32.	15.	7.	5.	3.	4.	0.	5.
241	2.	2.	6.	5.	4.	7.	2.	1.	6.	5.
251	3.	7.	0.	4.	8.	6.	4.	7.	5.	5.
261	9.	14.	11.	7.	9.	16.	13.	25.	30.	36.
271	20.	23.	14.	11.	12.	15.	13.	24.	15.	14.
281	23.	24.	25.	38.	46.	65.	81.	79.	89.	67.
291	33.	15.	9.	14.	10.	18.	12.	18.	17.	12.
301	16.	23.	24.	17.	29.	12.	15.	16.	8.	21.
311	13.	18.	8.	18.	24.	15.	12.	17.	10.	17.
321	27.	22.	19.	33.	33.	27.	32.	42.	39.	34.
331	46.	48.	50.	52.	54.	75.	69.	77.	80.	90.
341	84.	108.	113.	131.	151.	182.	194.	234.	280.	331.
351	437.	536.	641.	646.	578.	520.	533.	564.	690.	825.
361	1122.	1449.	1620.	1537.	1044.	396.	106.	11.	5.	2.
371	1.	2.	1.	0.	1.	1.	1.	1.	0.	1.
381	0.	1.	1.	0.	0.	0.	0.	1.	2.	0.
391	0.	0.	0.	1.	1.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.	2.
421	0.	0.	1.	3.	0.	1.	0.	0.	0.	1.
431	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
441	1.	0.	0.	0.	0.	0.	0.	1.	0.	0.
451	0.	1.	2.	1.	1.	0.	1.	0.	0.	2.
461	1.	0.	1.	0.	2.	0.	1.	0.	1.	5.
471	2.	2.	3.	3.	4.	6.	10.	4.	4.	3.
481	3.	1.	1.	0.	1.	0.	0.	0.	0.	0.
491	0.	0.	0.	1.	0.	0.	1.	1.	0.	1.
511	0.	0.								

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2963-SAM-PU
File ID: 9a9164.CNF

Counted on: 11/19/98 @ 2:57
Detector: AEA9
Geometry number: 1
Count time: 28805. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1179.3	1179.3	360.895	360.895	10.000	2.579	5.000	1.048
2	21.4	21.4	303.678	303.286	70.000	1.000	35.000	0.100
3?	19.4	19.4	299.769	298.826	6.000	0.000	3.000	2.765
4	43.9	43.9	285.769	285.673	8.000	2.102	4.000	1.733
5?	11.3	11.3	268.855	268.345	14.000	3.280	7.000	2.970
6	44.6	44.6	227.163	227.144	10.000	2.582	5.000	1.513

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err	d/m	Activity
			Exp.	Obs. Diff.	c/m	@95		uCi/ea
1	Pu236	0.866	5.755	5.759 -0.040	32.59	1.6	140.6	0.633E-04
	Cm243		5.779	5.759 0.020			188.7	0.850E-04
2	Pu238	0.046	5.487	5.494 -0.007	1.72	7.0	10.1	0.455E-05
	Am241		5.479	5.494 -0.015			7.7	0.349E-05
3		????		5.474	0.00	1000.		
4	Th228	0.021	5.400	5.413 -0.013	0.79	16.4	4.7	0.213E-05
5		????		5.333	0.23	42.6		
6	Pu239	0.027	5.147	5.144 0.003	1.00	10.1	4.2	0.190E-05
	Pu240		5.144	5.144 0.000			4.2	0.190E-05
Totals:			0.960	<--valid peaks only-->		36.11		

DETECTOR CALIBRATION

Energy (MEV) = 4.099 + (0.0046)*Channel
Energy range (MeV): 4.099 TO 6.454
Efficiency = 0.2366 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18065.0	100.000
Smoothed	18065.0	100.000
Composite fit	17445.7	96.572
Residuals	4958 619.3	3.428

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5802.8

6
 2.6
 2
 2
 2
 5
 5
 4
 2.4.
 4.
 2
 2
 ..
 .1.
1
1
 .....1....
 .....
1
 ...

Raw Data Dump for AEA Spectrum: 9a9164.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	1.	1.	0.	0.	1.	1.	0.	0.	0.	0.
31	0.	0.	0.	0.	0.	0.	0.	2.	0.	1.
41	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	1.	0.	0.	0.	1.	0.	1.	0.	0.
61	1.	1.	1.	1.	0.	0.	0.	0.	0.	0.
71	1.	0.	1.	0.	0.	0.	0.	0.	0.	2.
81	1.	1.	0.	1.	1.	2.	0.	1.	2.	0.
91	0.	1.	0.	0.	0.	1.	0.	0.	0.	1.
101	0.	1.	1.	0.	0.	1.	1.	1.	2.	0.
111	0.	0.	0.	1.	0.	1.	0.	2.	1.	1.
121	0.	0.	1.	0.	2.	0.	1.	1.	1.	1.
131	3.	3.	0.	0.	2.	0.	0.	1.	0.	0.
141	0.	1.	0.	0.	1.	1.	2.	2.	0.	0.
151	1.	2.	2.	0.	2.	0.	1.	1.	0.	0.
161	1.	0.	2.	0.	1.	2.	0.	2.	1.	0.
171	1.	1.	1.	2.	1.	5.	0.	0.	1.	1.
181	0.	2.	0.	1.	1.	2.	1.	2.	1.	1.
191	2.	0.	3.	2.	1.	6.	2.	5.	3.	2.
201	7.	1.	2.	4.	4.	3.	4.	10.	6.	4.
211	10.	2.	13.	8.	9.	21.	13.	13.	12.	23.
221	14.	23.	24.	37.	40.	45.	59.	64.	38.	27.
231	16.	2.	5.	4.	1.	3.	7.	2.	5.	0.
241	4.	7.	6.	11.	3.	6.	2.	2.	4.	7.
251	4.	5.	6.	11.	6.	6.	11.	7.	8.	3.
261	14.	8.	16.	12.	15.	18.	18.	25.	25.	20.
271	18.	13.	19.	20.	15.	13.	20.	18.	26.	26.
281	26.	39.	48.	49.	66.	75.	61.	49.	26.	12.
291	11.	18.	19.	22.	19.	11.	17.	17.	27.	27.
301	12.	19.	28.	22.	19.	22.	19.	24.	14.	15.
311	18.	19.	26.	27.	26.	28.	28.	29.	26.	33.
321	44.	30.	23.	44.	41.	41.	38.	42.	32.	44.
331	50.	63.	74.	70.	80.	85.	86.	89.	98.	104.
341	125.	164.	173.	191.	192.	242.	310.	331.	404.	501.
351	612.	656.	624.	582.	533.	548.	681.	772.	1005.	1300.
361	1502.	1411.	1015.	499.	138.	29.	2.	0.	1.	1.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
401	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	1.	0.	0.	0.	0.	1.	1.
421	1.	0.	0.	1.	0.	1.	0.	0.	0.	0.
431	1.	0.	0.	0.	0.	0.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	0.	0.	0.	1.	0.	0.	1.	1.
461	0.	1.	1.	0.	2.	0.	0.	0.	0.	1.
471	1.	3.	3.	4.	4.	8.	4.	0.	2.	1.
481	2.	0.	1.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2963-DUP-PU
 File ID: 10a1037.CNF

Counted on: 11/19/98 @ 2:57
 Detector: AEA10
 Geometry number: 1
 Count time: 28800. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1157.5	1157.5	361.805	361.805	10.000	2.786	5.000	1.170
2	18.4	18.4	304.162	303.950	56.000	1.000	28.000	0.100
3	32.0	32.0	287.331	287.202	10.000	2.292	5.000	2.031
4?	9.9	9.9	271.019	270.367	10.000	3.197	5.000	1.856
5	37.2	37.2	229.109	229.085	8.000	3.199	4.000	1.604

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.880	5.755	5.748	0.0070.01	31.80	1.6	138.1	0.622E-04
2	Pu238	0.041	5.487	5.482	0.0050.00	1.48	7.5	8.7	0.394E-05
	Am241		5.479	5.482	-0.003			6.7	0.301E-05
3	Th228	0.016	5.400	5.405	-0.0050.01	0.58	20.3	3.5	0.156E-05
4	????			5.327		0.23	39.5		
5	Pu239	0.026	5.147	5.137	0.0100.01	0.95	10.3	4.0	0.182E-05
	Pu240		5.144	5.137	0.007			4.0	0.182E-05

Totals: 0.964 <--valid peaks only--> 34.81

DETECTOR CALIBRATION

Energy (MEV) = 4.083 + (0.0046)*Channel
 Energy range (MeV): 4.083 TO 6.439
 Efficiency = 0.2350 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17339.0	100.000
Smoothed	17339.0	100.000
Composite fit	16819.2	97.002
Residuals	519.8	2.998

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 5878.4

5
25.
5
2
2
4
4
4
23.
3.
2
2
.
1.
...1
.....1
.....1..
..... 1
.....1
1....

Raw Data Dump for AEA Spectrum: 10a1037.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	1.	0.
31	0.	0.	0.	0.	0.	0.	0.	0.	0.
41	0.	1.	0.	0.	0.	0.	0.	1.	0.
51	0.	1.	1.	0.	0.	0.	0.	0.	0.
61	0.	1.	0.	0.	2.	1.	0.	1.	0.
71	0.	0.	0.	0.	1.	1.	0.	0.	0.
81	1.	1.	0.	0.	1.	0.	0.	2.	0.
91	0.	0.	1.	0.	0.	1.	0.	1.	0.
101	1.	0.	0.	0.	0.	0.	0.	0.	0.
111	0.	0.	1.	1.	0.	1.	0.	0.	2.
121	0.	0.	0.	0.	0.	0.	1.	0.	0.
131	0.	1.	0.	0.	1.	1.	0.	1.	0.
141	1.	2.	0.	0.	4.	2.	2.	1.	0.
151	0.	1.	0.	1.	3.	0.	3.	2.	1.
161	2.	1.	1.	0.	2.	3.	0.	2.	2.
171	1.	8.	2.	0.	1.	1.	1.	0.	1.
181	3.	0.	1.	4.	1.	0.	5.	2.	3.
191	2.	2.	3.	1.	6.	1.	3.	2.	1.
201	2.	1.	2.	2.	2.	5.	8.	4.	5.
211	4.	6.	3.	6.	11.	11.	7.	25.	12.
221	16.	18.	15.	25.	24.	32.	40.	41.	54.
231	36.	26.	11.	5.	4.	7.	6.	2.	7.
241	4.	3.	4.	6.	4.	3.	2.	2.	4.
251	3.	4.	2.	4.	3.	6.	5.	8.	8.
261	9.	11.	8.	13.	16.	11.	18.	17.	15.
271	22.	19.	14.	16.	15.	14.	9.	9.	7.
281	27.	20.	20.	38.	42.	38.	52.	61.	33.
291	17.	14.	14.	9.	15.	16.	12.	12.	12.
301	21.	14.	27.	13.	26.	19.	16.	10.	8.
311	9.	24.	26.	20.	20.	12.	17.	28.	18.
321	19.	27.	32.	24.	27.	31.	34.	40.	38.
331	55.	51.	56.	47.	55.	64.	67.	82.	83.
341	110.	110.	134.	149.	145.	173.	209.	268.	349.
351	518.	579.	593.	616.	532.	543.	572.	712.	814.
361	1318.	1434.	1365.	983.	510.	161.	37.	5.	0.
371	0.	1.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	1.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	1.	0.	0.	0.	0.	1.	0.	1.	0.
431	1.	1.	1.	0.	0.	1.	0.	0.	1.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	1.	0.	0.	0.	0.	0.	0.	0.	1.
461	0.	2.	0.	0.	0.	1.	1.	2.	2.
471	2.	4.	2.	0.	2.	6.	5.	3.	2.
481	1.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2964-SAM-PU
 File ID: 11a1134.CNF

Counted on: 11/19/98 @ 2:58
 Detector: AEA11
 Geometry number: 1
 Count time: 28807. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1193.3	1193.3	363.961	363.961	10.000	2.383	5.000	1.008
2	18.4	18.4	306.091	303.925	74.000	1.000	37.000	0.100
3	42.0	42.0	289.618	289.541	8.000	1.838	4.000	1.712
4?	10.8	10.8	272.195	271.806	12.000	3.306	6.000	12.125
5	54.1	54.1	231.176	231.158	12.000	3.505	6.000	1.790

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count	%err	d/m	Activity uCi/ea
		Exp.	Obs.	Diff.	FWHM	Rate c/m		
1	Pu236	0.859	5.755	5.750	0.0050.01	31.69	1.6	136.2 0.613E-04
2	Pu238	0.040	5.487	5.474	0.0130.00	1.48	7.5	8.7 0.391E-05
	Am241		5.479	5.474	0.005			6.6 0.299E-05
3	Th228	0.019	5.400	5.403	-0.0080.01	0.70	17.7	4.1 0.186E-05
4	????		5.576			0.20	47.4	
5	Pu239	0.038	5.147	5.139	0.0080.02	1.40	8.2	5.9 0.266E-05
	Pu240		5.144	5.139	0.005			5.9 0.266E-05

Totals: 0.956 <--valid peaks only--> 35.27

DETECTOR CALIBRATION

Energy (MEV) = 4.076 + (0.0046)*Channel
 Energy Range (MeV): 4.076 TO 6.431
 Efficiency = 0.2375 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17722.0	100.000
Smoothed	17722.0	100.000
Composite fit	17028.5	96.087
Residuals	693.5	3.913

4964 Analyzed by: _____

```

5
2.5
5
2
2
2
2
4
4
3.
3.
2
2
.
1.
.1.
.....1.
.....1.
.....1
.....1
.....1
.....1

```

Raw Data Dump for AEA Spectrum: 11a1134.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	0.	1.	1.
21	0.	0.	0.	0.	0.	0.	0.	1.	0.
31	1.	0.	0.	0.	0.	0.	1.	0.	0.
41	0.	0.	0.	0.	0.	0.	0.	1.	0.
51	0.	0.	0.	0.	0.	0.	0.	0.	0.
61	0.	0.	3.	2.	0.	0.	0.	0.	0.
71	0.	0.	0.	1.	0.	1.	0.	1.	0.
81	0.	0.	0.	0.	0.	0.	0.	0.	1.
91	0.	0.	1.	0.	0.	0.	0.	0.	0.
101	0.	1.	0.	0.	1.	0.	1.	0.	0.
111	0.	1.	0.	0.	1.	0.	0.	0.	1.
121	0.	1.	0.	0.	1.	0.	3.	1.	0.
131	0.	0.	0.	0.	3.	1.	1.	0.	0.
141	1.	0.	0.	0.	1.	0.	0.	1.	0.
151	1.	1.	0.	3.	0.	1.	3.	0.	0.
161	2.	0.	0.	0.	2.	2.	0.	2.	0.
171	1.	0.	0.	1.	4.	1.	0.	2.	0.
181	0.	2.	1.	1.	3.	3.	3.	3.	2.
191	1.	1.	5.	2.	3.	2.	3.	4.	2.
201	5.	6.	3.	3.	6.	2.	7.	3.	6.
211	7.	6.	5.	9.	14.	9.	15.	8.	9.
221	15.	22.	21.	25.	24.	36.	41.	40.	57.
231	71.	53.	69.	36.	23.	13.	5.	1.	4.
241	7.	4.	1.	4.	5.	4.	7.	3.	0.
251	3.	0.	5.	5.	2.	5.	3.	4.	5.
261	4.	4.	4.	17.	7.	9.	9.	22.	15.
271	21.	20.	27.	14.	13.	19.	12.	14.	10.
281	15.	20.	21.	17.	28.	41.	43.	51.	69.
291	67.	41.	18.	3.	18.	10.	10.	11.	15.
301	16.	18.	18.	27.	15.	16.	21.	21.	19.
311	15.	14.	17.	23.	10.	18.	21.	19.	19.
321	27.	22.	33.	17.	30.	33.	30.	29.	30.
331	44.	53.	46.	67.	53.	63.	68.	83.	80.
341	84.	95.	130.	117.	149.	156.	157.	196.	228.
351	301.	432.	497.	500.	701.	609.	511.	496.	517.
361	800.	1046.	1232.	1527.	1475.	1064.	449.	107.	11.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	0.	0.	0.	0.	0.	1.	0.	1.	0.
431	1.	0.	1.	1.	0.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	0.	1.	0.	1.	0.
461	0.	1.	2.	1.	0.	2.	0.	2.	1.
471	3.	1.	2.	1.	4.	3.	6.	4.	6.
481	3.	1.	2.	1.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T2964-DUP-PU
 File ID: 12a1210.CNF

Counted on: 11/19/98 @ 2:58
 Detector: AEA12
 Geometry number: 1
 Count time: 28808. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1280.2	1280.2	360.350	360.358	8.000	2.378	4.000	1.056
2	22.1	22.1	302.142	302.045	56.000	1.000	28.000	0.100
3	56.7	56.7	285.151	285.089	10.000	1.775	5.000	1.942
4?	11.8	11.8	267.959	267.538	10.000	2.072	5.000	1.356
5	60.5	60.5	226.901	226.901	10.000	2.855	5.000	1.732

PEAK RESULTS
 Peak Filter Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.					
1	Pu236	0.856	5.755	5.755	0.01	32.95	1.6	147.0	0.662E-04
2	Pu238	0.046	5.487	5.487	0.00	1.78	6.8	10.8	0.487E-05
	Am241		5.479	5.479	0.01			8.3	0.373E-05
3	Th228	0.023	5.400	5.400	0.01	0.88	15.2	5.4	0.243E-05
4	????		5.755	5.755		0.24	39.8		
5	Pu239	0.036	5.147	5.147	0.01	1.37	8.4	6.0	0.269E-05
	Pu240		5.144	5.144	0.00			6.0	0.269E-05
Totals:			0.960	<--valid peaks only-->		36.98			

DETECTOR CALIBRATION

Energy (MEV) = 0.100 + (0.0046)*Channel
 Energy (keV) = 4.100 TO 6.455
 Efficiency = 0.2287 CPM/DPM
 (Data reduction compression factor: 1.)

COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18484.0	100.000
Smoothed	18483.8	99.999
Composite fit	17866.7	96.661
Residuals	617.3	3.339

4967 Analyzed by: _____

Spectrum 12a1210.CNF
 1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 6396.3

```

.5
2.5.
2
2
2
4
4
3
2.3.
2
2
.
.
1.
...1
.....1
.....1.....
.....
.....1.
..

```

Raw Data Dump for AEA Spectrum: 12a1210.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	1.	0.
21	0.	0.	0.	0.	0.	0.	0.	0.	1.
31	0.	0.	1.	0.	0.	0.	0.	0.	1.
41	0.	0.	0.	1.	0.	0.	0.	0.	1.
51	0.	1.	0.	0.	0.	0.	0.	0.	0.
61	0.	0.	0.	1.	0.	0.	0.	0.	1.
71	0.	2.	0.	0.	1.	0.	0.	0.	2.
81	0.	0.	0.	0.	0.	1.	0.	0.	0.
91	2.	0.	0.	0.	0.	0.	3.	0.	1.
101	0.	0.	0.	0.	0.	0.	1.	0.	1.
111	0.	0.	1.	0.	0.	0.	0.	0.	0.
121	0.	1.	1.	1.	0.	0.	0.	0.	2.
131	0.	0.	0.	0.	1.	0.	0.	1.	3.
141	0.	1.	1.	1.	2.	1.	0.	1.	2.
151	0.	1.	1.	1.	2.	1.	1.	2.	0.
161	0.	0.	1.	2.	2.	2.	0.	3.	1.
171	4.	0.	0.	0.	0.	2.	1.	4.	1.
181	1.	3.	2.	0.	0.	1.	4.	1.	2.
191	4.	3.	7.	0.	5.	4.	5.	4.	2.
201	4.	3.	3.	2.	2.	4.	7.	9.	7.
211	2.	6.	16.	10.	12.	22.	12.	20.	21.
221	30.	21.	29.	5.	53.	78.	72.	80.	60.
231	14.	8.	6.	0.	2.	3.	5.	2.	2.
241	4.	3.	6.	0.	9.	4.	9.	5.	6.
251	8.	8.	5.	1.	10.	8.	8.	13.	12.
261	11.	4.	10.	1.	10.	14.	31.	20.	15.
271	15.	12.	14.	8.	11.	16.	17.	15.	29.
281	40.	49.	51.	70.	91.	87.	60.	42.	12.
291	13.	12.	10.	1.	10.	15.	13.	12.	15.
301	30.	28.	24.	2.	16.	14.	10.	20.	11.
311	11.	21.	16.	1.	21.	16.	36.	26.	23.
321	27.	32.	24.	3.	40.	29.	40.	39.	36.
331	50.	55.	61.	6.	67.	77.	70.	86.	89.
341	113.	139.	140.	10.	100.	240.	307.	383.	472.
351	728.	640.	540.	40.	530.	635.	756.	937.	1265.
361	1727.	1366.	690.	20.	20.	2.	0.	1.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	1.	0.
411	0.	0.	0.	0.	0.	0.	1.	0.	0.
421	0.	0.	0.	0.	0.	0.	0.	1.	0.
431	0.	0.	0.	0.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	0.	0.	0.	1.	0.	0.	0.
461	0.	1.	1.	0.	0.	0.	1.	1.	2.
471	5.	3.	0.	0.	0.	9.	2.	5.	0.
481	0.	1.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.	0.	0.	0.	0.	0.	0.	0.

LABCORE Completed Worklist Report for Worklist# 26839

Analyst: crj Instrument: AB14 Book#

Method: Rev/Mod

Worklist Comment: BY-105 FOR @PU23901 RTS

Seq Type	Sample#	RA	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@PU23901 PU23901 LIQUID		1.12E-04	1.13E-4	106.250	% Recovery
1 STD	0		@PU23901 PU23901E LIQUID		1.00	1.98E+00	1.980	% Ct Error
1 STD	0		@PU23901 PU23901T LIQUID		100	7.89E+01	78.900	% Recovery
2 BLNK	0		@PU23901 PU23901 LIQUID		1	<2.99E-5		uCi/mL
2 BLNK	0		@PU23901 PU23901T LIQUID		100	8.85E+01	88.500	% Recovery
2 BLNK	0		@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000	uCi/mL
3 SAMPLE	S98T003006	0	@PU23901 PU23901 LIQUID		N/A	<3.60E-05	3.60E-005	uCi/mL
3 SAMPLE	S98T003006	0	@PU23901 PU23901T LIQUID		N/A	7.80E+01		% Recovery
3 SAMPLE	S98T003006	0	@PU23901 PU23901E LIQUID		N/A	1.27E+01		% Ct. Error
4 DUP	S98T003006	0	@PU23901 PU23901 LIQUID		<3.60E-5	<3.57E-5		RPD
4 DUP	S98T003006	0	@PU23901 PU23901T LIQUID		100	7.76E+01	77.600	% Recovery
4 DUP	S98T003006	0	@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000	% Ct Error
5 SAMPLE	S98T003007	0	@PU23901 PU23901 LIQUID		N/A	<3.01E-05	3.01E-005	uCi/mL
5 SAMPLE	S98T003007	0	@PU23901 PU23901T LIQUID		N/A	8.67E+01		% Recovery
5 SAMPLE	S98T003007	0	@PU23901 PU23901E LIQUID		N/A	1.00E+02		% Ct. Error
6 DUP	S98T003007	0	@PU23901 PU23901 LIQUID		<3.01E-5	<3.43E-5		RPD
6 DUP	S98T003007	0	@PU23901 PU23901T LIQUID		100	8.18E+01	81.800	% Recovery
6 DUP	S98T003007	0	@PU23901 PU23901E LIQUID		1.00	1.44E+01	14.400	% Ct Error
7 SAMPLE	S98T003008	0	@PU23901 PU23901 LIQUID		N/A	<3.29E-05	3.29E-005	uCi/mL
7 SAMPLE	S98T003008	0	@PU23901 PU23901T LIQUID		N/A	8.14E+01		% Recovery
7 SAMPLE	S98T003008	0	@PU23901 PU23901E LIQUID		N/A	1.42E+01		% Ct. Error
8 DUP	S98T003008	0	@PU23901 PU23901 LIQUID		<3.29E-5	<3.44E-5		RPD
8 DUP	S98T003008	0	@PU23901 PU23901T LIQUID		100	8.15E+01	81.500	% Recovery
8 DUP	S98T003008	0	@PU23901 PU23901E LIQUID		1.00	1.18E+01	11.800	% Ct Error
9 SAMPLE	S98T003009	0	@PU23901 PU23901 LIQUID		N/A	<3.48E-05	3.48E-005	uCi/mL
9 SAMPLE	S98T003009	0	@PU23901 PU23901T LIQUID		N/A	8.08E+01		% Recovery
9 SAMPLE	S98T003009	0	@PU23901 PU23901E LIQUID		N/A	1.00E+02		% Ct. Error
10 DUP	S98T003009	0	@PU23901 PU23901 LIQUID		<3.48E-5	<3.46E-5		RPD
10 DUP	S98T003009	0	@PU23901 PU23901T LIQUID		100	7.93E+01	79.300	% Recovery
10 DUP	S98T003009	0	@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000	% Ct Error

Final page for worklist# 26839

Analyst Signature Date Analyst Signature Date

John Relyea 9 Dec 98
Reviewer Signature Date

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 26839

Analyst: CRS Instrument: PU01 #14 Book# 28B57

Method: LA-953-104 Rev/Mod B-0

Worklist Comment: BY-105 FOR @PU23901 RTS

S	Type	Sample#	R	A	Test	Matrix	Group#	Project
1	STD				@PU23901	LIQUID		
2	BLNK				@PU23901	LIQUID		
3	SAMPLE	S98T003006 0			@PU23901	LIQUID	98000510	BY-105 (2)
		Analytes Requested: PU23901 , PU23901E, PU23901T						
4	DUP	S98T003006 0			@PU23901	LIQUID		
5	SAMPLE	S98T003007 0			@PU23901	LIQUID	98000510	BY-105 (2)
		Analytes Requested: PU23901 , PU23901E, PU23901T						
6	DUP	S98T003007 0			@PU23901	LIQUID		
7	SAMPLE	S98T003008 0			@PU23901	LIQUID	98000510	BY-105 (2)
		Analytes Requested: PU23901 , PU23901E, PU23901T						
8	DUP	S98T003008 0			@PU23901	LIQUID		
9	SAMPLE	S98T003009 0			@PU23901	LIQUID	98000510	BY-105 (2)
		Analytes Requested: PU23901 , PU23901E, PU23901T						
10	DUP	S98T003009 0			@PU23901	LIQUID		

Final page for worklist # 26839

Analyst Signature

Date

Analyst Signature

Date

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**LIQUID**

				STD
Type	DATE COUNTED	DEC-08-98	Pu 236 AEA FRAC (C236)	0.362
STD	SAMPLE VOLUME in mL	SS	Pu 238 AEA FRAC (C238)	0.013
Work Unit	SAMPLE DILUTION FACTOR	DF	Pu 239 AEA FRAC (C239)	0.524
26839	TRACER VOLUME in mL	SPKV	TOTAL AT COUNTS	3487
Test Code	DIGEST DILUTION FACTOR	DDF	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	Pu 236 cpm	34.410
LIQUID	EFFICIENCY FACTOR	EFF	Pu 238 cpm	1.240
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	49.760
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Remn	PU-236 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 μ Ci/L	1.1937E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		
Sample Prep	STANDARD BOOK NO	28B57		
N/A	STANDARD VALUE in μ Ci/mL	1.120E-04		
Sample #				
WL26839-STD				
Instrument Code	Decay Time = Date Counted - Tracer Preparation Date			
WB27807	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]			
Prepared By	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236*100/Pu-236 Decay Corr'd Value * SPKV			
MCB	Pu 239/240 μ Ci/L = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(2220000 dpm/ μ Ci)]			
Chemist	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)			
JFR	Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-238 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]			
Analyst	Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100			
CRJ				
Date Complete				
12/09/98	Pu 239/240 μ Ci/mL	1.19E-04	DETECTION LEVELS in μ Ci/mL	
Analysis Date	Relative Counting Error	= 2.0%		
12/08/98				
01:50 PM			Pu 239/240	8.66E-06
Sample Point				
BY-105	Pu 236 Tracer Recovery	= 78.9%		

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98
STANDARD.WB1 REV 1.0 943128ML			

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**L****BLNK**

TYPE	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C238)	0.934
BLNK	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.000
WORK LIST	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.000
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1516
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cp (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	PU 238 cpm	38.330
LIQUID	EFFICIENCY FACTOR	EFF	0.2910 PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.00	AEA COUNT TIME	480
ReRun	PU-238 DECAY CORR'D VALUE (dpm/mL)	1830.70	Pu 239/240 $\mu\text{Ci/L}$ =	< 2.994E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.00		

Sample Prep

N/A

Sample #

Decay Time = Date Counted - Tracer Preparation Date

WL 26839-BLNK

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $^{\text{[e to the power of]}(-\ln 2 * \text{Decay Time}/1040.95)}$

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg)*C236*100/(Pu-236 Decay Corr'd Value * SPKV*EFF)

WB27807

Pu 239/240 $\mu\text{Ci/L}$ = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μCi)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value *SPKV *Pu 236 Tracer Recovery / 100)

MCB

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μCi)(D g/L)(SS)]

Chemists

Relative Counting Error = Square Root of $\{[(1/(\text{Pu 238 cpm} * \text{min})) + (1 / (\text{Pu 238 or 239/240 cpm} * \text{min}))]\} * 1.96 * 100$

JFR

Analyst

CRJ

Date Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Point

BY-105

Pu 239/240 $\mu\text{Ci/mL}$ < 2.99E-05

Relative Counting Error = 100.0%

DETECTION

LEVELS

in $\mu\text{Ci/mL}$

Pu 239/240

2.99E-05

Pu 238

2.99E-05

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 238 $\mu\text{Ci/mL}$ < 2.99E-05

Relative Counting Error = 100.0%

Pu 236 Tracer Recovery = 88.5%

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

BLANK.WB1 REV 1.0

44128ML

WORKBOOK PAGE: SAM3

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /**

				SAMPLE
TYPE	DATE COUNTED	DEC-98-98	PU 236 AEA FRAC (C236)	0.881
SAMPLE	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.013
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1417
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Main	DETECTOR NUMBER	14	PU 236 cpm	33.790
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.500
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
ReRun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.6003E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T003006

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date/Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Port

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/09/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.60E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
	Relative Counting Error	= 12.7%	
12/08/98	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
01:50 PM	Pu 238 $\mu\text{Ci/mL}$	< 3.60E-05	3.60E-05
	Relative Counting Error	= 100.0%	Pu 238
BY-105	Pu 236 Tracer Recovery	= 78.0%	3.60E-05

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: DUP4

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

Type	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C236)	0.893
DUP	SAMPLE VOLUME in mL	SS	PU 238 AEA FRAC (C238)	0.020
Worklist	SAMPLE DILUTION FACTOR	DF	PU 239 AEA FRAC (C239)	0.000
26839	TRACER VOLUME in mL	SPKV	TOTAL AT COUNTS	1391
Test Code	DIGEST DILUTION FACTOR	DDF	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	PU 238 cpm	38.950
LIQUID	EFFICIENCY FACTOR	EFF	PU 238 cpm	0.880
	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
Batch Number	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
98005443	PU-236 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.5698E-02
Refund	PU-238 TRACER VALUE (dpm/mL)	0.000		
0				
Sample Prep				
N/A				

Sample Prep

N/A

Sample #

S98T003006

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) * (Pu 236 Decay Corr'd Value) * (SPKV) * (1000mL/L) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) * (DF) * (DDF) * (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μCi) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/09/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.57E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
Analysis Date	Relative Counting Error	= 100.0%	
12/08/98			Pu 239/240
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		3.57E-05
01:50 PM	Pu 238 $\mu\text{Ci/mL}$	< 3.57E-05	Pu 238
Sample Point	Relative Counting Error	= 9.6%	3.57E-05
BY-105	Pu 236 Tracer Recovery	= 77.6%	

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: SAM5

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

Type	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C236)	0.947
SAMPLE	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.000
WORKLIST	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.000
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1466
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	PU 238 cpm	35.610
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Return	PU-238 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 μ Ci/L =	< 3.0117E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T003007

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times$ [e to the power of $\{(-\ln 2 \times \text{Decay Time}/1040.95)\}$]

Instrument Code

WB27807

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times$ C236 $\times$ 100 / (Pu-236 Decay Corr'd Value $\times$ SPKV $\times$ EFF)

Prepared By

MCB

Pu 239/240 μ Ci/L = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μ Ci)]Pu 238 dpm = [(Total AT Counts / TC) - Bkg $\times$ 1/EFF $\times$ C238] - (Pu-238 Tracer Value $\times$ SPKV $\times$ Pu 236 Tracer Recovery / 100)

Chemist

JFR

Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]

Analyst

CRJ

Relative Counting Error = Square Root of $\{[1/(\text{Pu 238 cpm} \times \text{min})] + [1/(\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

Date Complete

12/09/98

Analysis Date

12/08/98

Pu 239/240 μ Ci/mL

Relative Counting Error = < 3.01E-05

DETECTION

LEVELS

in μ Ci/mL

Analysis Time

01:50 PM

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 238 μ Ci/mL < 3.01E-05

Sample Point

BY-105

Relative Counting Error = 100.0%

Pu 236 Tracer Recovery = 86.7%

Pu 239/240

3.01E-05

Pu 238

3.01E-05

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: DUP6

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID /

DUP

Type	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C236)	0.883
DUP	SAMPLE VOLUME in mL SS	0.100	PU 238 AEA FRAC (C238)	0.044
Worklist	SAMPLE DILUTION FACTOR DF	1.000	PU 239 AEA FRAC (C239)	0.010
26839	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	1482
Test Code	DIGEST DILUTION FACTOR DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	PU 238 cpm	34.190
LIQUID	EFFICIENCY FACTOR EFF	0.291	PU 238 cpm	1.710
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.390
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Refund	PU-238 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 μ Ci/L =	< 3.4266E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		
Sample Prep				
N/A				

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {-(ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL) * (DF) * (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μ Ci)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 μ Ci/L = [(Pu 238 dpm) * (DF) * (DDF) / (1000 mL)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Chemist	JFR		
Analyst	CRJ		
Date Complete	12/09/98	Pu 239/240 μ Ci/mL	< 3.43E-05
Analysis Date	12/08/98	Relative Counting Error	= 14.4%
Analysis Time	01:50 PM	Pu 238 μ Ci/mL	4.11E-05
Sample Point	BY-105	Relative Counting Error	= 7.0%
		Pu 236 Tracer Recovery	= 81.8%
		DETECTION LEVELS in μ Ci/mL	Pu 239/240 3.43E-05
			Pu 238 3.43E-05

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

WORKBOOK PAGE: SAM7

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

TYPE	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C236)	0.925
SAMPLE	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.031
WORKLIST	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.011
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1408
TEST CODE	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
MATRIX	DETECTOR NUMBER	14	PU 238 cpm	33.450
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	1.110
BATCH NUMBER	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.400
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
REFUR	PU-238 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.2869E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T003008

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of ((-ln2 * Decay Time/1040.95))]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL) * (DF) / (DDF) / ((C236) * (D g/L) / (2220000 dpm/ μCi))

Pu 238 dpm = ((Total AT Counts / TC) - Bkg * 1/EFF * C238) - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = ((Pu 238 dpm) * (DF) * (DDF) / (1000 mL)) / ((Pu-236 Tracer Recovery / 100) / (2220000 dpm/ μCi) * (D g/L) * (SS))

Relative Counting Error = Square Root of ((1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))) * 1.96 * 100

12/09/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.29E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
12/08/98	Relative Counting Error	= 14.2%	
01:50 PM	NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
	Pu 238 $\mu\text{Ci/mL}$	< 3.29E-05	3.29E-05
	Relative Counting Error	= 8.6%	Pu 238
	Pu 236 Tracer Recovery	= 81.4%	3.29E-05

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4978

WORKBOOK PAGE: DUP8

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

Type	DATE COUNTED	DEC-08-98	PU 238 AEA FRAC (C236)	0.884
DUP	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.058
Work List	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.015
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1475
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	PU 236 cpm	34.680
LIQUID	EFFICIENCY FACTOR	EFF	0.291 PU 238 cpm	2.260
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.580
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Return	PU-236 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.4351E-02
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T003008

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Completed

12/09/98

Analysis Date

12/09/98

Analysis Time

01:50 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times [e \text{ to the power of } \{-\ln 2 \times \text{Decay Time}/1040.95\}]$ Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times \text{C236} \times 100 / (\text{Pu-236 Decay Corr'd Value} \times \text{SPKV} \times \text{EFF})$ Pu 239/240 $\mu\text{Ci/L}$ = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μCi)]Pu 238 dpm = [(Total AT Counts / TC) - Bkg $\times 1/\text{EFF} \times \text{C238}$] - (Pu-238 Tracer Value $\times \text{SPKV} \times \text{Pu 236 Tracer Recovery} / 100$)Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μCi)(D g/L)(SS)]Relative Counting Error = Square Root of $\{[1/(\text{Pu 236 cpm} \times \text{min})] + [1/(\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

12/09/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.44E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
12/09/98	Relative Counting Error =	11.8%	
01:50 PM	Pu 238 $\mu\text{Ci/mL}$	5.41E-05	Pu 239/240
BY-105	Relative Counting Error =	6.1%	3.44E-05
	Pu 236 Tracer Recovery =	81.5%	Pu 238
			3.44E-05

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98
SAMPLE.WB1 REV 1.0	943128ML		

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LIQUID /

	DATE COUNTED		DEC-08-98	PU 238 AEA FRAC (C236)	0.880
SAMPLE	SAMPLE VOLUME in mL	SS	0.100	PU 238 AEA FRAC (C238)	0.045
WORKLIST	SAMPLE DILUTION FACTOR	DF	1.000	PU 239 AEA FRAC (C239)	0.000
26839	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1470
TEST CODE	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO		145B43	BACKGROUND in cpm (Bkg)	0.070
MAIN	DETECTOR NUMBER		14	PU 238 cpm	33.150
LIQUID	EFFICIENCY FACTOR	EFF	0.291	PU 238 cpm	1.690
BATCH NUMBER	TRACER PREPARATION DATE		01/19/98	PU 239 cpm	0.000
98005443	TRACER PREPARATION VALUE (dpm/mL)		2270.000	AEA COUNT TIME	480
RECUR	PU-238 DECAY CORR'D VALUE (dpm/mL)		1830.704	Pu 239/240 $\mu\text{Ci/L}$ =	< 3.4782E-02
0	PU-238 TRACER VALUE (dpm/mL)		0.000		

Sample Prep

N/A

Sample #

S98T003009

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Pu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) / (DDF) / [(C236) * (SS) * (D g/L) * (2220000 dpm/ μCi)]

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) * (DF) * (DDF) / (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μCi) * (D g/L) * (SS)]

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

12/09/98	Pu 239/240 $\mu\text{Ci/mL}$	< 3.48E-05	DETECTION LEVELS in $\mu\text{Ci/mL}$
Analysis Date	Relative Counting Error	= 100.0%	
12/08/98			Pu 239/240
Analysis Time	Pu 238 $\mu\text{Ci/mL}$	4.22E-05	3.48E-05
01:50 PM	Relative Counting Error	= 7.1%	Pu 238
Sample Point	Pu 236 Tracer Recovery	= 80.8%	3.48E-05
BY-105			

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	JFR	Date:	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: DUP10

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

Type	DATE COUNTED	DEC-08-98	Pu 236 AEA FRAC (C236)	0.902
DUP	SAMPLE VOLUME in mL	SS	0.100 Pu 238 AEA FRAC (C238)	0.027
Work List	SAMPLE DILUTION FACTOR	DF	1.000 Pu 239 AEA FRAC (C239)	0.000
26839	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	1407
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.070
Matrix	DETECTOR NUMBER	14	Pu 236 cpm	34.860
LIQUID	EFFICIENCY FACTOR	EFF	0.291 Pu 238 cpm	1.080
Batch Number	TRACER PREPARATION DATE	01/19/98	Pu 239 cpm	0.000
98005443	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Result	PU-236 DECAY CORR'D VALUE (dpm/mL)	1830.704	Pu 239/240 μ Ci/L =	< 3.4591E-02
0	PU-236 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample #

S98T003009

Instrument Code

WB27807

Prepared By

MCB

Chemist

JFR

Analyst

CRJ

Date Complete

12/09/98

Analysis Date

12/08/98

Analysis Time

01:50 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times$ [e to the power of $\{-\ln 2 \times \text{Decay Time} / 1040.95\}$]Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times$ C236 $\times$ 100 / (Pu-236 Decay Corr'd Value $\times$ SPKV $\times$ EFF)Pu 239/240 μ Ci/L = (C239) (Pu 236 Decay Corr'd Value) (SPKV) (1000mL/L) (DF) (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μ Ci)]Pu 238 dpm = [(Total AT Counts / TC) - Bkg $\times$ 1/EFF $\times$ C238] - (Pu-238 Tracer Value $\times$ SPKV $\times$ Pu 236 Tracer Recovery / 100)Pu 238 μ Ci/L = [(Pu 238 dpm) (DF) (DDF) (1000mL/L)] / [(Pu-236 Tracer Recovery / 100) (2220000 dpm/ μ Ci) (D g/L) (SS)]Relative Counting Error = Square Root of $\{[1 / (\text{Pu 236 cpm} \times \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

12/09/98	Pu 239/240 μ Ci/mL	< 3.46E-05	DETECTION LEVELS
Analysis Date	Relative Counting Error	= 100.0%	in μ Ci/mL
12/08/98			Pu 239/240
Analysis Time	NOTE: Pu 238 Result is a LESS THAN Value.		3.46E-05
01:50 PM	Pu 238 μ Ci/mL	< 3.46E-05	Pu 238
Sample Point	Relative Counting Error	= 8.8%	3.46E-05
BY-105	Pu 236 Tracer Recovery	= 79.3%	

Analyst:	CRJ	Date:	09-Dec-98
Signature of Chemist:	<i>John Relyea</i>	JFR	9 Dec 98

SAMPLE.WB1 REV 1.0

943128ML

4981

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL26839-STD-PU
File ID: 13a1388.CNF

Counted on: 12/ 8/98 @20: 1
Detector: AEA13
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	987.9	987.9	363.957	363.957	10.000	2.714	5.000	0.793
2	40.5	40.5	303.701	303.312	32.000	9.551	16.000	16.301
3	61.6	61.6	289.330	289.051	12.000	3.260	6.000	0.787
4	1699.5	1699.5	230.949	230.948	10.000	2.971	5.000	1.149

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.	c/m			
1	Pu236	0.362	5.755	5.758	-.0030	34.41	1.5	141.5	0.637E-04
2	Am241	0.013	5.479	5.479	0.0000	1.24	10.6	5.3	0.240E-05
	Pu238		5.487	5.479	0.008			7.0	0.314E-05
3	Th228	0.027	5.400	5.413	-.0130	2.53	6.2	14.4	0.646E-05
4	Pu239	0.524	5.147	5.146	0.0010	49.76	1.3	200.5	0.903E-04
	Pu240		5.144	5.146	-.002			200.5	0.903E-04
Totals:			0.926	<--valid peaks only-->			87.94		

DETECTOR CALIBRATION

Energy (MEV) = 4.083 + (0.0046)*Channel
Energy range (MeV): 4.083 TO 6.439
Efficiency = 0.2482 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	45580.0	100.000
Smoothed	45580.0	100.000
Composite fit	42212.6	92.612
Residuals	3367.4	7.388

Analyzed by:

NEW

Spectrum 13a1388.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 10497.7

```

.
.
.
.
.
4..
.4...
....4...
.....4...
.....4
.....4
.....4
.....4
.
3
3
3
.3
3.
2
2
1
.1
...1
.....1
.....1
.....1
.....1
.....1
..1..

```


Raw Data Dump for AEA Spectrum: 13a1388.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	1.	1.	3.	3.	1.	2.
21	1.	1.	1.	2.	1.	2.	3.	3.	0.
31	4.	3.	3.	5.	2.	1.	0.	1.	2.
41	0.	2.	3.	7.	1.	3.	1.	1.	0.
51	3.	2.	0.	3.	3.	2.	0.	3.	2.
61	4.	2.	4.	0.	2.	2.	5.	2.	3.
71	5.	4.	4.	3.	3.	0.	4.	3.	3.
81	7.	4.	5.	4.	3.	3.	4.	5.	3.
91	6.	7.	8.	4.	7.	5.	5.	4.	6.
101	7.	5.	6.	1.	7.	8.	3.	7.	4.
111	3.	6.	9.	8.	12.	7.	5.	5.	11.
121	6.	6.	9.	5.	5.	17.	10.	8.	10.
131	13.	8.	13.	3.	12.	9.	11.	15.	12.
141	12.	15.	12.	14.	18.	12.	13.	19.	17.
151	10.	13.	19.	26.	13.	19.	22.	22.	30.
161	19.	27.	24.	31.	33.	24.	27.	27.	22.
171	43.	36.	29.	38.	47.	41.	41.	55.	38.
181	54.	48.	58.	56.	54.	61.	73.	60.	73.
191	77.	79.	91.	87.	91.	113.	116.	122.	115.
201	146.	135.	162.	148.	191.	203.	181.	213.	226.
211	284.	310.	290.	359.	392.	445.	479.	515.	596.
221	718.	772.	763.	785.	836.	965.	1096.	1339.	1585.
231	2013.	2043.	1595.	924.	344.	91.	22.	5.	6.
241	6.	5.	5.	8.	7.	11.	10.	3.	5.
251	13.	4.	6.	13.	8.	8.	7.	12.	9.
261	14.	13.	17.	21.	17.	16.	15.	19.	26.
271	46.	24.	31.	35.	18.	19.	32.	17.	24.
281	36.	40.	32.	39.	43.	63.	59.	72.	77.
291	77.	48.	32.	33.	17.	28.	28.	27.	29.
301	36.	44.	36.	43.	51.	32.	30.	20.	27.
311	33.	23.	29.	38.	24.	44.	31.	40.	26.
321	45.	46.	43.	54.	62.	55.	79.	76.	65.
331	81.	83.	79.	99.	99.	94.	119.	155.	114.
341	149.	169.	202.	196.	203.	234.	245.	251.	295.
351	380.	448.	573.	658.	644.	671.	594.	535.	588.
361	740.	875.	958.	1191.	1209.	929.	447.	144.	27.
371	1.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	1.	0.	0.	0.	1.	0.	1.	0.	0.
431	0.	0.	1.	0.	0.	1.	0.	0.	0.
441	0.	0.	0.	1.	0.	0.	2.	0.	0.
451	0.	1.	0.	0.	0.	1.	0.	3.	1.
461	0.	0.	0.	0.	3.	0.	3.	1.	0.
471	0.	1.	3.	3.	2.	5.	3.	6.	2.
481	2.	3.	2.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL26839-BLK-PU
 File ID: 14a1474.CNF

Counted on: 12/ 8/98 @20: 2
 Detector: AEA14
 Geometry number: 1
 Count time: 28805. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	773.5	773.5	361.384	361.384	10.000	3.232	5.000	0.610
2	47.0	47.0	286.768	286.469	14.000	4.937	7.000	0.730

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count	%err	Activity	
		Exp.	Obs.	Diff.	FWHM	Rate	c/m	@95	d/m
1	Pu236	0.934	5.755	5.755	0.0000.01	38.33	1.4	161.0	0.725E-04
2	Th228	0.072	5.400	5.410	-0.0100.02	2.96	6.3	17.1	0.772E-05
Totals:		1.006	<--valid peaks only-->				41.29		

DETECTOR CALIBRATION

Energy(MEV) = 4.092 + (0.0046)*Channel
 Energy range (MeV): 4.092 TO 6.448
 Efficiency = 0.2430 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	19706.0	100.000
Smoothed	19706.0	100.000
Composite fit	19822.0	100.589
Residuals	-116.0	-0.589

Analyzed by:

NEW

Spectrum 14a1474.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,.., etc

Display Max.: 4787.4

```

2
2
2.
2.
.2.
1.2.
1...2.
.2..
..1.
....1
.....1
..... 1
..... 1
..... 1
.....1.
.....1
.....1
1...

```

Raw Data Dump for AEA Spectrum: 14a1474.CNF

HNF-1663 REV. 0

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	1.	0.	1.
21	1.	2.	1.	1.	0.	0.	0.	1.	0.
31	1.	0.	1.	0.	0.	2.	3.	0.	1.
41	0.	0.	0.	0.	1.	2.	0.	2.	0.
51	0.	0.	0.	1.	1.	1.	0.	0.	1.
61	1.	1.	0.	1.	0.	0.	2.	1.	0.
71	3.	0.	0.	3.	0.	0.	0.	0.	2.
81	1.	1.	0.	1.	1.	5.	0.	2.	1.
91	1.	2.	1.	0.	0.	1.	0.	1.	1.
101	1.	0.	1.	0.	2.	0.	3.	4.	1.
111	3.	0.	2.	0.	0.	2.	0.	2.	1.
121	3.	2.	3.	1.	1.	1.	4.	1.	1.
131	2.	5.	0.	1.	1.	1.	5.	0.	2.
141	3.	2.	0.	4.	1.	1.	0.	2.	3.
151	1.	1.	2.	3.	0.	2.	5.	4.	3.
161	5.	1.	0.	6.	2.	8.	1.	1.	2.
171	6.	1.	0.	2.	4.	1.	3.	2.	4.
181	3.	3.	3.	5.	4.	5.	4.	2.	3.
191	1.	4.	3.	4.	3.	2.	5.	4.	5.
201	9.	0.	4.	3.	6.	5.	3.	6.	2.
211	6.	8.	3.	6.	6.	6.	0.	9.	6.
221	5.	5.	3.	7.	5.	4.	4.	12.	6.
231	8.	7.	7.	5.	10.	6.	5.	9.	14.
241	11.	13.	13.	18.	16.	14.	9.	11.	13.
251	15.	26.	23.	18.	19.	14.	23.	13.	11.
261	18.	30.	30.	24.	31.	23.	29.	34.	35.
271	37.	57.	27.	37.	29.	46.	27.	40.	45.
281	37.	55.	35.	49.	69.	73.	73.	62.	63.
291	41.	25.	28.	30.	40.	27.	30.	31.	30.
301	43.	47.	49.	40.	38.	49.	50.	52.	53.
311	46.	50.	55.	50.	63.	75.	73.	66.	71.
321	78.	107.	95.	91.	105.	107.	107.	107.	133.
331	149.	130.	150.	178.	169.	183.	206.	218.	223.
341	253.	301.	303.	354.	304.	369.	351.	389.	438.
351	546.	616.	662.	611.	554.	493.	598.	607.	650.
361	825.	922.	842.	615.	313.	81.	11.	1.	1.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	1.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	1.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	1.
421	1.	0.	0.	0.	0.	1.	0.	1.	0.
431	2.	2.	0.	3.	0.	0.	2.	1.	0.
441	0.	0.	2.	0.	0.	1.	0.	0.	1.
451	1.	0.	1.	1.	1.	1.	2.	1.	1.
461	2.	0.	0.	1.	2.	2.	1.	0.	2.
471	2.	1.	2.	7.	8.	5.	2.	3.	2.
481	2.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	1.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3006-SAM-PU
File ID: 15a1554.CNF

Counted on: 12/ 8/98 @20: 3
Detector: AEA15
Geometry number: 1
Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	923.4	923.4	362.006	362.006	10.000	3.021	5.000	0.842
2	54.1	54.1	287.227	287.156	16.000	11.150	8.000	6.125
3	20.8	20.8	270.406	269.421	12.000	5.184	6.000	0.965
4	10.9	10.9	230.293	230.072	16.000	8.141	8.000	2.042

PEAK RESULTS

Peak Error Limit: 30%

Peak ID Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
	Exp.	Obs.	Diff.	FWHM				
1 Pu236	0.881	5.755	5.753	0.0020.01	33.79	1.5	142.9	0.644E-04
2 Th228	0.051	5.400	5.408	-.0080.05	1.96	7.0	11.4	0.516E-05
3	0.029		5.327	0.02	1.10	11.0	4.6	0.205E-05
4 Pu239	0.013	5.147	5.146	0.0010.04	0.50	15.8	2.1	0.942E-06
Pu240		5.144	5.146	-.002			2.1	0.942E-06

Totals: 0.974 <--valid peaks only--> 37.35

DETECTOR CALIBRATION

Energy(MEV) = 4.087 + (0.0046)*Channel
Energy range (MeV): 4.087 TO 6.443
Efficiency = 0.2412 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18417.0	100.000
Smoothed	18417.0	100.000
Composite fit	17930.8	97.360
Residuals	486.2	2.640

Analyzed by:

NEW

Spectrum 15a1554.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 4927.6

```

4
4
4
4
4
3
3
3
.3
32
....2
...2
2..
1..
..1.
.....1.
.....1
..... 1
.....1..
..... 1
.....1
..1...

```

Raw Data Dump for AEA Spectrum: 15a1554.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	1.	1.	0.
21	0.	1.	1.	0.	0.	0.	1.	0.	1.
31	0.	0.	1.	0.	2.	1.	1.	0.	0.
41	1.	0.	2.	1.	1.	0.	0.	0.	1.
51	0.	0.	0.	0.	0.	0.	2.	0.	1.
61	1.	0.	0.	1.	0.	0.	0.	0.	1.
71	0.	0.	0.	1.	0.	1.	0.	1.	0.
81	0.	2.	0.	1.	1.	1.	0.	1.	2.
91	0.	2.	0.	0.	1.	1.	2.	2.	1.
101	0.	1.	0.	0.	2.	0.	0.	2.	1.
111	1.	2.	0.	1.	0.	1.	1.	1.	0.
121	2.	0.	1.	1.	2.	1.	2.	0.	2.
131	4.	1.	0.	1.	1.	1.	0.	3.	3.
141	1.	3.	2.	3.	2.	1.	3.	3.	0.
151	3.	1.	4.	1.	1.	1.	2.	2.	2.
161	1.	2.	1.	2.	3.	4.	2.	0.	0.
171	2.	3.	3.	2.	2.	2.	4.	3.	2.
181	5.	2.	1.	3.	5.	2.	0.	3.	0.
191	0.	0.	3.	4.	4.	2.	6.	3.	4.
201	3.	6.	9.	3.	4.	1.	5.	3.	6.
211	3.	3.	5.	4.	5.	4.	9.	3.	12.
221	5.	4.	12.	10.	6.	11.	4.	11.	6.
231	13.	8.	10.	8.	8.	12.	6.	11.	8.
241	11.	8.	9.	7.	9.	7.	6.	5.	9.
251	11.	7.	7.	18.	11.	12.	19.	12.	10.
261	21.	9.	23.	21.	21.	26.	17.	25.	26.
271	30.	21.	26.	25.	25.	23.	28.	22.	28.
281	32.	26.	41.	58.	44.	75.	60.	64.	58.
291	25.	24.	14.	28.	25.	26.	29.	12.	25.
301	29.	25.	28.	32.	38.	30.	32.	18.	33.
311	29.	35.	44.	25.	26.	44.	55.	40.	38.
321	55.	54.	53.	54.	60.	85.	80.	101.	81.
331	107.	102.	97.	103.	116.	132.	143.	131.	163.
341	179.	185.	230.	228.	253.	296.	280.	343.	389.
351	513.	543.	641.	590.	561.	549.	550.	636.	644.
361	949.	1098.	1111.	874.	516.	224.	55.	12.	1.
371	0.	0.	0.	1.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	1.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	2.	0.	0.
421	0.	0.	0.	1.	0.	1.	0.	2.	2.
431	1.	1.	3.	1.	1.	2.	1.	1.	1.
441	0.	0.	0.	1.	0.	0.	0.	0.	0.
451	0.	0.	2.	0.	0.	0.	1.	0.	0.
461	1.	0.	0.	0.	0.	3.	1.	1.	1.
471	2.	3.	6.	5.	9.	5.	5.	3.	7.
481	1.	0.	2.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3006-DUP-PU
File ID: 16a1666.CNF

Counted on: 12/ 8/98 @20: 4
Detector: AEA16
Geometry number: 1
Count time: 28803. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	938.5	938.5	361.567	361.567	10.000	3.091	5.000	0.729
2	18.7	18.7	304.784	302.426	14.000	19.059	7.000	8.098
3	47.6	47.6	287.464	287.111	10.000	3.190	5.000	0.836
4	15.5	15.5	268.212	267.323	12.000	3.376	6.000	0.524

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err @95	d/m	Activity uCi/ea	
			Exp.	Obs.	Diff.					
1	Pu236	0.893	5.755	5.751	0.0040	0.01	38.95	1.4	146.0	
2	Am241	0.020	5.479	5.479	0.0000	0.09	0.88	21.4	3.4	
	Pu238		5.487	5.479	0.008				4.5	
3	Th228	0.042	5.400	5.408	-0.0080	0.01	1.84	8.0	9.5	
4		0.021		5.317	0.02		0.91	13.4	3.3	
Totals:			0.976	<--valid peaks only-->			42.57			

DETECTOR CALIBRATION

Energy(MEV) = 4.088 + (0.0046)*Channel
Energy range (MeV): 4.088 TO 6.443
Efficiency = 0.2723 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	20938.0	100.000
Smoothed	20937.6	99.998
Composite fit	20437.1	97.608
Residuals	500.9	2.392

Analyzed by:

NEW

Spectrum 16a1666.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5391.9

4
4
4
4
4
4
4.
.3
2..3.
3..
2..
2.1.
2...1.
.....1
.....1
.....1
.....1.
..... 1
.....1
1...

Raw Data Dump for AEA Spectrum: 16a1666.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	1.	0.	1.
21	0.	2.	0.	1.	1.	1.	0.	1.	0.
31	1.	0.	1.	0.	1.	0.	1.	1.	0.
41	0.	1.	0.	1.	0.	1.	0.	0.	0.
51	0.	0.	0.	1.	1.	1.	1.	0.	1.
61	0.	0.	0.	0.	1.	2.	0.	1.	1.
71	0.	0.	0.	1.	0.	3.	1.	1.	0.
81	0.	3.	1.	1.	1.	0.	1.	2.	1.
91	1.	2.	3.	0.	0.	1.	2.	0.	2.
101	0.	2.	1.	1.	0.	0.	0.	3.	0.
111	2.	0.	2.	1.	3.	1.	0.	1.	2.
121	0.	1.	0.	2.	0.	1.	1.	1.	4.
131	1.	2.	4.	2.	0.	0.	3.	0.	3.
141	2.	1.	1.	2.	3.	4.	4.	2.	3.
151	3.	0.	0.	2.	4.	0.	3.	4.	2.
161	4.	1.	2.	2.	1.	2.	1.	2.	2.
171	2.	1.	3.	3.	5.	3.	5.	3.	2.
181	3.	0.	2.	0.	2.	4.	4.	2.	4.
191	1.	1.	6.	1.	1.	5.	4.	4.	5.
201	6.	7.	2.	6.	2.	4.	3.	7.	6.
211	4.	11.	5.	6.	16.	7.	7.	9.	8.
221	10.	6.	3.	7.	9.	16.	10.	13.	9.
231	16.	15.	4.	8.	7.	13.	8.	8.	7.
241	11.	7.	9.	10.	14.	2.	11.	16.	21.
251	15.	13.	7.	14.	20.	11.	21.	17.	19.
261	29.	19.	21.	16.	19.	30.	35.	38.	29.
271	31.	28.	16.	28.	32.	29.	22.	29.	31.
281	40.	43.	45.	60.	58.	55.	72.	56.	76.
291	28.	29.	22.	22.	25.	32.	31.	28.	28.
301	37.	42.	45.	39.	43.	44.	40.	30.	44.
311	36.	48.	49.	54.	58.	66.	73.	64.	63.
321	80.	76.	97.	88.	80.	98.	103.	120.	141.
331	140.	153.	161.	158.	149.	164.	197.	198.	253.
341	227.	258.	304.	305.	329.	350.	363.	425.	448.
351	564.	624.	697.	692.	626.	634.	590.	674.	775.
361	963.	1113.	1092.	764.	395.	132.	21.	2.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	1.	0.	0.	0.	0.
401	1.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	1.	0.	0.	1.	0.	2.
421	0.	0.	0.	0.	0.	0.	0.	0.	1.
431	0.	0.	1.	2.	1.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	1.	0.	0.	0.
451	1.	1.	1.	0.	1.	1.	2.	0.	0.
461	1.	2.	2.	0.	0.	1.	2.	2.	2.
471	1.	5.	2.	4.	5.	4.	9.	3.	4.
481	0.	3.	1.	0.	0.	1.	0.	0.	1.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T3007-SAM-PU
 File ID: 17a1708.CNF

Counted on: 12/ 8/98 @20: 5
 Detector: AEA17
 Geometry number: 1
 Count time: 28803. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	786.5	786.5	362.983	362.983	10.000	3.014	5.000	0.632
2	41.6	41.6	287.890	287.630	8.000	5.508	4.000	2.570
3	15.1	15.1	272.217	270.160	22.000	3.254	11.000	0.362

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.					
1	Pu236	0.947	5.755	5.764	-0.0090.01	35.61	1.5	148.8	0.670E-04	
	Cm243		5.779	5.764	0.015			199.8	0.900E-04	
2	Th228	0.034	5.400	5.417	-0.0170.03	1.27	10.3	7.3	0.329E-05	
3		0.031		5.337	0.01	1.16	10.3	4.8	0.215E-05	
Totals:			1.012	<--valid peaks only-->			38.04			

DETECTOR CALIBRATION

Energy (MeV) = 4.094 + (0.0046)*Channel
 Energy range (MeV): 4.094 TO 6.449
 Efficiency = 0.2441 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18048.0	100.000
Smoothed	18048.0	100.000
Composite fit	18259.2	101.170
Residuals	-211.2	-1.170

Analyzed by:

NEW

Spectrum 17a1708.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 4685.3

```

3
3
3
3
3
3
3
3
3
3.
3.
1.2.
12..
.1.
...1
.....1
..... 1
..... 1
..... 1
.....1.
.....1
.....1
...1...

```

Raw Data Dump for AEA Spectrum: 17a1708.CNF						HNF-1663 REV. 0				
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	
11	0.	0.	0.	0.	1.	1.	0.	0.	0.	
21	0.	0.	1.	0.	0.	0.	1.	0.	0.	
31	0.	0.	0.	0.	0.	0.	0.	0.	2.	
41	2.	0.	0.	0.	0.	0.	2.	1.	0.	
51	0.	0.	1.	0.	1.	1.	1.	1.	0.	
61	0.	0.	0.	0.	2.	2.	0.	0.	1.	
71	0.	1.	0.	0.	0.	0.	1.	0.	0.	
81	0.	3.	0.	0.	0.	0.	0.	0.	0.	
91	0.	1.	0.	1.	3.	1.	1.	0.	0.	
101	0.	4.	0.	1.	0.	0.	0.	0.	1.	
111	0.	0.	1.	1.	1.	2.	1.	1.	2.	
121	1.	2.	2.	0.	0.	2.	2.	1.	0.	
131	0.	2.	0.	0.	1.	1.	0.	4.	0.	
141	2.	2.	1.	0.	2.	0.	2.	1.	2.	
151	1.	1.	1.	0.	0.	2.	1.	1.	1.	
161	0.	3.	1.	1.	1.	1.	0.	1.	0.	
171	0.	3.	0.	0.	0.	2.	1.	0.	4.	
181	2.	1.	2.	4.	1.	0.	0.	2.	3.	
191	3.	5.	4.	4.	3.	0.	3.	6.	3.	
201	2.	4.	3.	2.	5.	6.	5.	1.	4.	
211	4.	4.	3.	8.	7.	8.	8.	7.	7.	
221	4.	11.	5.	9.	4.	10.	10.	8.	7.	
231	8.	9.	2.	7.	8.	6.	3.	11.	5.	
241	3.	6.	9.	9.	6.	7.	4.	3.	11.	
251	10.	22.	13.	7.	10.	9.	9.	16.	14.	
261	12.	14.	21.	16.	28.	20.	20.	13.	17.	
271	27.	20.	22.	23.	22.	19.	19.	18.	27.	
281	33.	32.	39.	39.	49.	59.	50.	48.	72.	
291	33.	25.	24.	20.	32.	26.	29.	28.	37.	
301	29.	29.	27.	26.	34.	47.	27.	39.	27.	
311	37.	45.	44.	51.	40.	51.	46.	52.	45.	
321	72.	64.	56.	71.	70.	98.	86.	94.	97.	
331	112.	127.	135.	133.	147.	154.	167.	151.	175.	
341	188.	241.	255.	267.	280.	299.	302.	355.	389.	
351	466.	453.	598.	622.	588.	542.	512.	491.	499.	
361	679.	779.	929.	921.	754.	461.	162.	37.	5.	
371	0.	1.	0.	0.	1.	0.	0.	0.	0.	
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	
411	0.	0.	0.	0.	1.	0.	0.	0.	0.	
421	0.	0.	0.	0.	2.	2.	0.	0.	0.	
431	0.	0.	0.	1.	1.	0.	1.	0.	1.	
441	0.	0.	0.	0.	0.	1.	0.	0.	0.	
451	0.	0.	2.	1.	1.	1.	2.	2.	0.	
461	0.	1.	1.	1.	1.	3.	0.	2.	2.	
471	0.	0.	1.	1.	0.	0.	4.	5.	3.	
481	0.	1.	2.	1.	0.	0.	0.	0.	0.	
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	
511	1.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3007-DUP-PU
File ID: 18a1881.CNF

Counted on: 12/ 8/98 @20: 6
Detector: AEA18
Geometry number: 1
Count time: 28803. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1087.0	1087.0	360.978	360.978	10.000	2.763	5.000	0.939
2	21.3	21.3	302.990	302.649	64.000	1.000	32.000	0.100
3	49.3	49.3	285.912	285.810	10.000	2.419	5.000	2.148
4?	13.0	13.0	269.494	269.019	16.000	3.540	8.000	1.900
5	10.9	10.9	227.815	227.738	10.000	4.258	5.000	1.318

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err c/m	Activity d/m	Activity uCi/ea
		Exp.	Obs.	Diff.	FWHM	@95		
1	Pu236	0.883	5.755	5.750	0.0050.01	34.19 1.5	148.1	0.667E-04
2	Pu238	0.044	5.487	5.482	0.0050.00	1.71 7.2	10.1	0.455E-05
	Am241		5.479	5.482	-0.003		7.7	0.349E-05
3	Th228	0.023	5.400	5.404	-0.0040.01	0.91 14.9	5.4	0.244E-05
4		????		5.327		0.33 31.8		
5	Pu239	0.010	5.147	5.137	0.0100.02	0.39 18.8	1.7	0.754E-06
	Pu240		5.144	5.137	0.007		1.7	0.754E-06

Totals: 0.961 <--valid peaks only--> 37.20

DETECTOR CALIBRATION

Energy (MEV) = 4.090 + (0.0046)*Channel
Energy range (MeV): 4.090 TO 6.445
Efficiency = 0.2356 CPM/DFM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18590.0	100.000
Smoothed	18590.0	100.000
Composite fit	18016.1	96.913
Residuals	573.9	3.087

Spectrum 18a1881.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5702.5

```

5
5
5
5
5
2
2
4
4
4.
2.3..
3.
2.
..
1..
..1..
.....1.
.....1
.....1...
.....1
.....1
.....1

```

Raw Data Dump for AEA Spectrum: 18a1881.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	1.	0.	0.	0.
21	2.	0.	0.	2.	0.	0.	0.	0.	0.
31	1.	0.	1.	0.	0.	0.	1.	0.	0.
41	0.	3.	0.	0.	0.	0.	1.	0.	0.
51	0.	1.	0.	0.	1.	1.	0.	0.	0.
61	1.	1.	1.	1.	1.	1.	0.	1.	0.
71	0.	2.	0.	1.	1.	1.	1.	1.	0.
81	0.	1.	0.	0.	1.	0.	1.	0.	1.
91	0.	0.	0.	0.	1.	0.	1.	0.	0.
101	0.	0.	0.	1.	2.	0.	0.	1.	2.
111	1.	0.	0.	0.	0.	0.	1.	1.	0.
121	0.	1.	2.	2.	1.	1.	1.	1.	1.
131	1.	2.	0.	1.	1.	1.	0.	0.	0.
141	3.	1.	0.	2.	2.	2.	0.	0.	1.
151	1.	1.	0.	1.	1.	1.	1.	4.	0.
161	2.	1.	2.	0.	1.	2.	0.	0.	4.
171	2.	0.	3.	1.	1.	0.	0.	1.	1.
181	1.	1.	1.	1.	2.	0.	5.	3.	1.
191	4.	5.	5.	3.	0.	2.	2.	3.	4.
201	3.	7.	3.	2.	2.	5.	6.	4.	8.
211	7.	5.	2.	5.	4.	9.	6.	5.	9.
221	10.	11.	12.	5.	8.	12.	17.	16.	17.
231	8.	8.	2.	7.	5.	6.	5.	5.	4.
241	8.	6.	4.	9.	3.	5.	8.	5.	8.
251	11.	7.	8.	6.	6.	16.	11.	12.	8.
261	9.	7.	13.	17.	17.	17.	17.	29.	26.
271	27.	14.	20.	23.	18.	14.	16.	24.	22.
281	34.	41.	49.	63.	80.	71.	80.	57.	38.
291	17.	13.	12.	16.	11.	21.	13.	15.	25.
301	24.	23.	30.	29.	21.	15.	18.	24.	23.
311	32.	28.	23.	25.	28.	29.	43.	36.	37.
321	38.	47.	36.	56.	47.	53.	76.	83.	66.
331	82.	82.	102.	73.	105.	122.	93.	135.	139.
341	180.	173.	220.	226.	265.	274.	329.	375.	449.
351	606.	635.	635.	566.	538.	587.	702.	778.	986.
361	1314.	1308.	1040.	538.	160.	39.	9.	1.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	1.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	2.	1.	1.
411	0.	0.	1.	1.	1.	0.	1.	0.	0.
421	1.	0.	1.	2.	0.	0.	0.	0.	1.
431	0.	0.	1.	2.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	2.
451	0.	0.	1.	1.	1.	0.	2.	0.	1.
461	1.	2.	4.	0.	1.	1.	4.	1.	2.
471	2.	6.	6.	1.	5.	4.	4.	4.	3.
481	2.	1.	1.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3008-SAM-PU
File ID: 19a1905.CNF

Counted on: 12/ 8/98 @20: 7
Detector: AEA19
Geometry number: 1
Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	951.5	951.5	362.393	362.393	8.000	2.809	4.000	0.816
2	16.8	16.8	305.276	304.299	16.000	38.423	8.000	12.662
3	25.4	25.4	288.271	287.981	6.000	1.922	3.000	0.700
4	12.2	12.2	229.611	229.597	10.000	5.169	5.000	1.978

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err @95	d/m	Activity uCi/ea	
			Exp.	Obs. Diff.	c/m				
1	Pu236	0.925	5.755	5.752	0.0030.01	33.45	1.5	142.5 0.642E-04	
2	Am241	0.031	5.479	5.484	-0.0050.18	1.11	19.6	4.9 0.221E-05	
	Pu238		5.487	5.484	0.003			6.4 0.289E-05	
3	Th228	0.020	5.400	5.409	-0.0090.01	0.72	13.7	4.2 0.191E-05	
4	Pu239	0.011	5.147	5.141	0.0060.02	0.40	14.3	1.7 0.758E-06	
	Pu240		5.144	5.141	0.003			1.7 0.758E-06	

Totals:		0.986	<--valid peaks only-->			35.67			

DETECTOR CALIBRATION

Energy (MEV) = 4.084 + (0.0046)*Channel
Energy range (MeV): 4.084 TO 6.440
Efficiency = 0.2395 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17363.0	100.000
Smoothed	17363.0	100.000
Composite fit	17123.9	98.623
Residuals	239.1	1.377

Analyzed by:

NEW

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5266.9

4
4
4
4

3

3

23

3.

2.

2.

21.

2...1.

.....1

..... 1

.....1.

..... 1

.....1

.1....

Raw Data Dump for AEA Spectrum: 19a1905.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	2.	1.	1.	2.
21	0.	0.	0.	1.	0.	1.	0.	0.	0.
31	1.	0.	1.	0.	0.	0.	1.	0.	0.
41	0.	0.	0.	1.	0.	0.	1.	0.	0.
51	0.	2.	0.	1.	0.	0.	0.	0.	1.
61	1.	1.	1.	0.	0.	1.	1.	1.	0.
71	1.	1.	0.	0.	0.	1.	0.	1.	0.
81	0.	0.	1.	0.	0.	0.	0.	1.	0.
91	1.	0.	0.	0.	0.	1.	0.	1.	0.
101	0.	0.	2.	0.	0.	0.	0.	1.	1.
111	0.	0.	1.	0.	1.	0.	0.	1.	1.
121	0.	0.	1.	1.	0.	1.	0.	0.	0.
131	0.	0.	2.	1.	2.	2.	0.	0.	0.
141	1.	0.	0.	2.	0.	0.	0.	2.	1.
151	1.	6.	1.	0.	0.	2.	0.	1.	0.
161	0.	1.	1.	1.	0.	2.	0.	0.	3.
171	1.	2.	1.	1.	2.	2.	1.	2.	2.
181	1.	3.	2.	2.	1.	0.	3.	0.	0.
191	1.	2.	3.	2.	2.	2.	5.	3.	1.
201	2.	1.	2.	3.	2.	2.	6.	7.	1.
211	7.	4.	4.	4.	1.	1.	2.	4.	4.
221	4.	4.	4.	8.	11.	7.	11.	10.	13.
231	15.	5.	10.	2.	3.	4.	4.	0.	4.
241	3.	3.	6.	5.	8.	2.	7.	4.	1.
251	2.	5.	3.	7.	5.	5.	9.	6.	5.
261	5.	9.	8.	14.	11.	9.	5.	11.	8.
271	15.	12.	18.	13.	12.	14.	11.	12.	19.
281	16.	22.	20.	28.	29.	28.	39.	33.	45.
291	15.	10.	15.	12.	21.	20.	27.	16.	19.
301	26.	20.	25.	23.	30.	31.	24.	20.	24.
311	32.	18.	30.	20.	29.	34.	38.	43.	38.
321	37.	44.	56.	62.	55.	76.	67.	82.	77.
331	95.	90.	104.	102.	100.	109.	142.	137.	136.
341	166.	174.	172.	207.	243.	237.	278.	328.	386.
351	491.	538.	597.	636.	603.	533.	522.	601.	663.
361	940.	1096.	1135.	1120.	647.	254.	57.	7.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	1.	0.	0.	0.	0.	0.
411	0.	0.	0.	1.	0.	0.	0.	0.	0.
421	1.	0.	1.	0.	1.	1.	1.	2.	0.
431	0.	0.	2.	0.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	0.	0.	0.	1.	0.	0.	0.
461	0.	0.	1.	0.	1.	1.	1.	3.	4.
471	1.	1.	2.	2.	2.	4.	4.	4.	1.
481	3.	0.	1.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3008-DUP-PU
File ID: 20a2072.CNF

Counted on: 12/ 8/98 @20: 8
Detector: AEA20
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	998.3	998.3	362.878	362.878	10.000	2.912	5.000	0.867
2	25.6	25.6	305.316	304.843	46.000	77.735	23.000	32.237
3	37.0	37.0	287.983	287.632	10.000	1.761	5.000	0.976
4	18.8	18.8	270.507	269.515	14.000	5.574	7.000	1.553
5	11.3	11.3	229.519	229.421	12.000	5.265	6.000	1.013

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.					
1	Pu236	0.884	5.755	5.766	-0.011	0.01	34.68	1.5	148.6	0.669E-04
	Cm243		5.779	5.766	0.013				199.5	0.898E-04
2	Am241	0.058	5.479	5.499	-0.020	0.36	2.26	16.7	10.1	0.455E-05
	Pu238		5.487	5.499	-0.012				13.2	0.593E-05
3	Th228	0.020	5.400	5.420	-0.020	0.01	0.78	14.5	4.6	0.209E-05
4		0.019		5.337		0.03	0.76	14.0	3.2	0.144E-05
5	Pu239	0.015	5.147	5.152	-0.005	0.02	0.58	12.5	2.4	0.110E-05
	Pu240		5.144	5.152	-0.008				2.4	0.110E-05

Totals: 0.996 <---valid peaks only---> 39.07

DETECTOR CALIBRATION

Energy (MeV) = 4.097 + (0.0046) * Channel

Energy range (MeV): 4.097 TO 6.452

Efficiency = 0.2382 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18829.0	100.000
Smoothed	18829.0	100.000
Composite fit	18754.4	99.604
Residuals	74.6	0.396

Spectrum 20a2072.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

5850.6

5
 5
 5
 5
 5
 5
 4
 4
 4
 4
 4.
 23.
 3..
 .2
 12
 21.
 2..1.
 2.....1
1
 .....1.
 1
1
 ...1...

Raw Data Dump for AEA Spectrum: 20a2072.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	1.	0.	1.	1.	1.	1.
21	0.	0.	0.	1.	0.	0.	1.	1.	0.	2.
31	1.	0.	0.	0.	0.	1.	0.	0.	0.	1.
41	0.	0.	0.	1.	1.	0.	0.	0.	2.	1.
51	1.	0.	2.	0.	0.	0.	0.	0.	1.	1.
61	2.	1.	0.	1.	0.	0.	0.	0.	1.	0.
71	0.	0.	0.	2.	0.	2.	2.	1.	0.	1.
81	0.	0.	0.	1.	1.	2.	0.	0.	1.	1.
91	0.	0.	0.	0.	0.	0.	0.	2.	2.	1.
101	1.	0.	0.	0.	1.	0.	0.	2.	0.	0.
111	0.	1.	1.	1.	0.	2.	1.	2.	0.	0.
121	0.	1.	0.	0.	1.	0.	0.	1.	1.	0.
131	1.	2.	0.	1.	1.	1.	0.	0.	0.	2.
141	0.	3.	0.	1.	2.	0.	3.	1.	1.	1.
151	2.	1.	3.	1.	1.	0.	1.	3.	0.	1.
161	1.	2.	2.	2.	1.	1.	2.	1.	3.	2.
171	2.	1.	1.	0.	1.	0.	1.	2.	0.	1.
181	1.	1.	1.	2.	2.	2.	1.	0.	1.	1.
191	4.	1.	2.	6.	2.	1.	2.	3.	2.	6.
201	3.	3.	5.	2.	6.	5.	4.	6.	5.	5.
211	7.	6.	4.	10.	2.	7.	5.	3.	8.	4.
221	10.	9.	9.	6.	6.	8.	10.	14.	10.	16.
231	14.	8.	4.	10.	3.	1.	1.	7.	6.	3.
241	6.	5.	6.	7.	2.	6.	5.	10.	5.	8.
251	8.	10.	12.	5.	8.	12.	7.	14.	11.	6.
261	8.	6.	13.	14.	17.	20.	15.	22.	30.	33.
271	20.	22.	25.	26.	16.	23.	14.	22.	28.	16.
281	24.	26.	32.	30.	54.	43.	73.	51.	71.	53.
291	27.	14.	22.	16.	18.	12.	24.	19.	25.	25.
301	26.	28.	26.	27.	44.	44.	23.	27.	22.	31.
311	28.	38.	40.	35.	36.	43.	46.	34.	45.	56.
321	53.	51.	54.	61.	71.	64.	68.	75.	72.	82.
331	82.	93.	107.	106.	114.	121.	120.	121.	160.	168.
341	168.	202.	196.	232.	229.	259.	288.	366.	328.	362.
351	433.	538.	612.	656.	661.	548.	557.	570.	657.	744.
361	905.	1076.	1181.	1190.	929.	495.	171.	28.	0.	1.
371	1.	1.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	1.	0.	1.	0.	0.
421	0.	0.	1.	0.	1.	1.	0.	0.	1.	0.
431	0.	1.	1.	0.	1.	0.	0.	0.	0.	0.
441	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	1.	0.	0.	1.	0.	0.	1.	0.
461	1.	0.	1.	1.	0.	2.	0.	2.	1.	1.
471	3.	6.	3.	1.	6.	8.	4.	10.	9.	3.
481	1.	2.	3.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	1.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T3009-SAM-PU
 File ID: 21a2185.CNF

Counted on: 12/ 8/98 @20: 9
 Detector: AEA21
 Geometry number: 1
 Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1077.8	1077.8	364.775	364.775	10.000	2.504	5.000	0.859
2	20.9	20.9	307.188	306.428	56.000	1.000	28.000	0.100
3	44.3	44.3	290.100	290.001	10.000	2.184	5.000	1.207
4?	11.7	11.7	272.220	271.415	8.000	2.450	4.000	1.067
5?	9.7	9.7	231.910	231.759	10.000	2.630	5.000	1.033

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.				
1	Pu236	0.880	5.755	5.760	-0.0050.01	33.15	1.6	140.6	0.633E-04
	Cm243		5.779	5.760	0.019			188.7	0.850E-04
2	Pu238	0.045	5.487	5.492	-0.0050.00	1.69	7.3	9.7	0.438E-05
	Am241		5.479	5.492	-0.013			7.5	0.336E-05
3	Th228	0.026	5.400	5.416	-0.0160.01	0.99	13.9	5.8	0.260E-05
4	????			5.331		0.31	33.1		
5	????			5.148		0.27	24.3		
Totals:			0.951	<--valid peaks only-->			35.82		

DETECTOR CALIBRATION

Energy(MEV) = 4.082 + (0.0046)*Channel
 Energy range (MeV): 4.082 TO 6.437
 Efficiency = 0.2406 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18078.0	100.000
Smoothed	18078.0	100.000
Composite fit	17476.6	96.673
Residuals	601.4	3.327

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 6762.8

```

5
5
5
5
2
2
4
4
4
3.
23.
2
2
..
.1.
...1.
.....1
.....1
.....1.
..... 1
.....1...

```


Raw Data Dump for AEA Spectrum: 21a2185.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
31	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
41	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
51	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
61	1.	0.	0.	0.	1.	0.	1.	1.	0.	0.
71	0.	0.	0.	0.	1.	0.	0.	1.	1.	1.
81	1.	1.	0.	2.	1.	0.	3.	0.	0.	0.
91	0.	0.	0.	2.	0.	2.	1.	0.	1.	0.
101	0.	0.	1.	2.	0.	1.	1.	2.	0.	0.
111	0.	1.	0.	0.	1.	0.	0.	0.	0.	0.
121	1.	0.	0.	1.	1.	2.	0.	0.	1.	1.
131	1.	1.	0.	0.	1.	1.	1.	2.	0.	1.
141	0.	1.	0.	0.	1.	0.	1.	0.	0.	1.
151	2.	0.	0.	2.	2.	3.	1.	0.	1.	3.
161	1.	0.	0.	1.	1.	0.	1.	0.	0.	1.
171	1.	1.	1.	3.	1.	0.	0.	2.	3.	2.
181	2.	0.	0.	1.	0.	4.	4.	4.	3.	1.
191	4.	2.	4.	1.	0.	4.	1.	1.	2.	3.
201	3.	4.	3.	1.	3.	1.	1.	3.	1.	5.
211	3.	0.	2.	3.	1.	3.	6.	5.	5.	7.
221	7.	9.	5.	8.	10.	4.	8.	8.	10.	11.
231	12.	18.	10.	12.	8.	6.	4.	2.	2.	4.
241	2.	4.	6.	3.	5.	6.	7.	5.	3.	3.
251	5.	6.	1.	5.	11.	7.	6.	13.	13.	13.
261	5.	10.	14.	11.	16.	15.	21.	11.	19.	25.
271	23.	28.	21.	31.	10.	15.	20.	16.	23.	19.
281	32.	22.	29.	36.	25.	35.	42.	50.	67.	72.
291	67.	59.	35.	12.	21.	20.	14.	17.	14.	14.
301	20.	15.	21.	22.	25.	24.	34.	21.	24.	24.
311	23.	15.	33.	26.	24.	24.	33.	25.	36.	30.
321	29.	29.	32.	50.	41.	39.	48.	53.	55.	58.
331	69.	66.	75.	76.	71.	89.	96.	104.	121.	110.
341	133.	142.	156.	123.	168.	186.	229.	236.	243.	296.
351	332.	329.	463.	568.	646.	684.	615.	589.	518.	563.
361	594.	754.	955.	1190.	1317.	1346.	895.	342.	66.	10.
371	3.	1.	1.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
421	0.	0.	0.	1.	1.	3.	0.	0.	1.	1.
431	0.	1.	0.	1.	0.	1.	0.	0.	0.	0.
441	1.	1.	0.	1.	0.	0.	1.	0.	0.	0.
451	1.	1.	1.	0.	0.	1.	0.	0.	0.	0.
461	0.	0.	0.	0.	1.	0.	1.	0.	0.	1.
471	0.	1.	4.	0.	2.	2.	4.	4.	3.	6.
481	3.	7.	4.	2.	2.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
 G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 S98T3009-DUP-PU
 File ID: 22a2271.CNF

Counted on: 12/ 8/98 @20:10
 Detector: AEA22
 Geometry number: 1
 Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	971.0	971.0	360.590	360.590	10.000	2.876	5.000	0.815
2	20.3	20.3	302.449	301.351	22.000	26.055	11.000	12.886
3	55.3	55.3	285.388	285.179	10.000	2.617	5.000	1.448
4	19.2	19.2	269.466	268.737	14.000	4.000	7.000	0.933
5?	10.9	10.9	227.205	227.112	10.000	3.948	5.000	1.450

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.902	5.755	5.756	-0.0010.01	34.86	1.5	151.5	0.682E-04
2	Am241	0.027	5.479	5.484	-0.0050.12	1.06	16.4	4.8	0.217E-05
	Pu238		5.487	5.484	0.003			6.3	0.283E-05
3	Th228	0.033	5.400	5.409	-0.0090.01	1.28	9.3	7.7	0.346E-05
4		0.021		5.334	0.02	0.83	12.8	3.5	0.159E-05
5		????		5.142		0.35	17.2		
-----						-----			
Totals:			0.984 <--valid peaks only-->			38.04			

DETECTOR CALIBRATION

Energy (MEV) = 4.098 + (0.0046) * Channel

Energy range (MeV): 4.098 TO 6.453

Efficiency = 0.2349 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18557.0	100.000
Smoothed	18557.0	100.000
Composite fit	18426.0	99.294
Residuals	131.0	0.706

5009 Analyzed by:

NEW

Spectrum 22a2271.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

5539.0

```

5
5
5
4
4
4
4
4.
4.
2..3.
3.
2.
2.
2.1
.....1
.....1
..... 1
.....1....
..... 1
.....1
...

```

Raw Data Dump for AEA Spectrum: 22a2271.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	2.	0.	0.	0.	0.	1.	0.	0.
31	0.	0.	0.	0.	0.	1.	0.	1.	0.	1.
41	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	0.	1.	0.	1.	2.
61	0.	0.	0.	0.	2.	1.	1.	0.	1.	0.
71	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
81	0.	0.	1.	0.	0.	1.	1.	0.	0.	1.
91	0.	1.	0.	0.	0.	0.	0.	1.	1.	1.
101	0.	1.	1.	0.	0.	0.	0.	0.	1.	1.
111	0.	1.	0.	1.	0.	1.	1.	0.	1.	2.
121	1.	0.	1.	1.	0.	0.	0.	0.	1.	0.
131	1.	0.	1.	0.	0.	1.	3.	0.	0.	1.
141	0.	1.	0.	0.	0.	2.	1.	1.	2.	0.
151	2.	0.	0.	0.	1.	1.	3.	1.	1.	0.
161	2.	0.	1.	1.	1.	1.	2.	1.	2.	2.
171	1.	1.	1.	1.	0.	1.	4.	2.	3.	1.
181	5.	1.	1.	1.	4.	0.	2.	2.	0.	0.
191	2.	3.	3.	1.	2.	2.	3.	4.	2.	3.
201	3.	4.	2.	6.	4.	2.	5.	1.	5.	3.
211	0.	4.	2.	8.	8.	2.	7.	7.	7.	4.
221	9.	2.	4.	4.	11.	8.	17.	20.	5.	7.
231	4.	4.	5.	5.	6.	8.	1.	7.	7.	6.
241	7.	6.	5.	13.	5.	9.	4.	5.	7.	8.
251	8.	7.	6.	7.	8.	10.	9.	11.	19.	22.
261	13.	12.	12.	15.	28.	21.	17.	30.	29.	28.
271	30.	14.	25.	15.	27.	16.	22.	28.	26.	25.
281	44.	42.	61.	52.	82.	75.	67.	37.	31.	16.
291	23.	14.	16.	25.	23.	22.	20.	19.	30.	26.
301	27.	40.	28.	29.	24.	29.	27.	27.	34.	30.
311	32.	36.	27.	26.	46.	42.	42.	38.	50.	51.
321	53.	55.	57.	67.	64.	67.	71.	74.	82.	110.
331	121.	105.	96.	106.	125.	142.	168.	163.	160.	188.
341	215.	220.	262.	273.	304.	324.	409.	405.	508.	564.
351	593.	657.	622.	542.	528.	636.	667.	789.	880.	1062.
361	1163.	1146.	774.	348.	107.	11.	4.	0.	0.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
391	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	1.	1.	0.	1.
411	0.	0.	0.	0.	1.	2.	0.	1.	1.	1.
421	2.	0.	1.	0.	2.	0.	0.	0.	0.	0.
431	1.	0.	1.	1.	0.	0.	0.	0.	0.	0.
441	1.	0.	0.	0.	0.	1.	0.	0.	0.	0.
451	0.	0.	0.	1.	0.	2.	0.	2.	0.	0.
461	0.	2.	1.	1.	4.	1.	1.	4.	3.	1.
471	4.	7.	5.	13.	4.	4.	3.	3.	0.	5.
481	3.	0.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

LABCORE Completed Worklist Report for Worklist# 27632

Analyst: rro Instrument: AB13 Book#: _____

Method: LA-953-104 Rev/Mod _____

Worklist Comment: BY-105 @PU23901 RLC

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0	@PU23901	PU23901	LIQUID	1.12E-04	1.25E-4	111.607	% Recovery
1 STD	0	@PU23901	PU23901T	LIQUID	100	8.27E+01	82.700	% Recovery
1 STD	0	@PU23901	PU23901E	LIQUID	1.00	1.93E+00	1.930	% Ct Error
2 BLNK	0	@PU23901	PU23901	LIQUID	1	<2.93E-5		uCi/mL
2 BLNK	0	@PU23901	PU23901T	LIQUID	100	8.82E+01	88.200	% Recovery
2 BLNK	0	@PU23901	PU23901E	LIQUID	1.00	1.00E+02	100.000	uCi/mL
3 SAMPLE	S98T003384	0	@PU23901	PU23901	LIQUID	N/A	1.30E-04	1.30E-004 uCi/mL
3 SAMPLE	S98T003384	0	@PU23901	PU23901T	LIQUID	N/A	2.27E+01	% Recovery
3 SAMPLE	S98T003384	0	@PU23901	PU23901E	LIQUID	N/A	9.34E+00	% Ct Error
4 DUP	S98T003384	0	@PU23901	PU23901	LIQUID	<1.30E-4	3.46E-4	RPD
4 DUP	S98T003384	0	@PU23901	PU23901T	LIQUID	100	1.83E+01	18.300 % Recovery
4 DUP	S98T003384	0	@PU23901	PU23901E	LIQUID	1.00	6.00E+00	6.000 % Ct Error

Final page for worklist# 27632

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

John Peljea 4/Jan 99
 Reviewer Signature _____ Date _____

LABCORE Data Entry Template for Worklist# 27632

Analyst: RRO Instrument: PU01 13 Book# 28B57

Method: LA-953-104 Rev/Mod B-0

Worklist Comment: BY-105 @PU23901 RLC

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@PU23901	LIQUID		
2 BLNK			@PU23901	LIQUID		
3 SAMPLE	S98T003384 0		@PU23901	LIQUID	98000437	BY-105 (2)
	Analytes Requested: PU23901, PU23901E, PU23901T					
4 DUP	S98T003384 0		@PU23901	LIQUID		

Final page for worklist # 27632

Robin R O'Deal 12-30-98
Analyst Signature Date

Sandra Z. Chom 12/31/98
Analyst Signature Date
HEWright 1/1/99

Data Entry Comments:

Sample exhausted by analysis of 10R 4/29/99

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)

LIQUID

STD

Type	DATE COUNTED	DEC-30-98	PU 236 AEA FRAC (C236)	0.372		
STD	SAMPLE VOLUME in mL	SS	1.000	PU 238 AEA FRAC (C238)	0.011	
Work List	SAMPLE DILUTION FACTOR	DF	1.000	PU 239 AEA FRAC (C239)	0.571	
27632	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	3811	
Test Code	DIGEST DILUTION FACTOR	DDF	1.000	AT COUNT TIME (MIN)	30	
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.100		
Matrix	DETECTOR NUMBER	13	PU 236 cpm	35.370		
LIQUID	EFFICIENCY FACTOR	EFF	0.317	PU 238 cpm	1.020	
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	54.230		
98006245	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480		
Recon	PU-236 DECAY CORR'D VALUE (dpm/mL)	1804.081	Pu 239/240 μ Ci/L	1.2474E-01		
0	PU-238 TRACER VALUE (dpm/mL)	0.000				
Sample Prep	STANDARD BOOK NO	28B57				
N/A	STANDARD VALUE in μ Ci/mL	1.120E-04				
Sample #						
WL27632						
Instrument Code	Decay Time = Date Counted - Tracer Preparation Date					
WB27810	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]					
Prepared By	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236*100/Pu-236 Decay Corr'd Value * SPKV					
SZC	Pu 239/240 μ Ci/L = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)/(1000mL/L)(DF)(DDF) / [(C236)(SS)(2220000 dpm/ μ Ci)]					
Chemist	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value *SPKV *Pu 236 Tracer Recovery / 100)					
JFR	Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery /100)(2220000 dpm/ μ Ci)(D g/L)(SS)]					
Analyst	Relative Counting Error = Square Root of [(1/(Pu 238 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100					
RRO						
Date Complete						
12/31/98	Pu 239/240 μ Ci/mL	1.25E-04	DETECTION LEVELS in μ Ci/mL Pu 239/240 7.93E-06			
Analysis Date	Relative Counting Error	= 1.9%				
12/30/98						
Analysis Time						
11:15 AM						
Sample Point						
BY-105 (2)	Pu 236 Tracer Recovery	= 82.7%				

Analyst:	RRO	Date:	31-Dec-98
Signature of Chemist:	JFR	Date:	4/3 Jan 99
STANDARD.WB1 REV 1.0	943128ML		87a

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

LI

BLNK

Type	DATE COUNTED	DEC-30-98	PU 236 AEA FRAC (C236)	0.943
BLNK	SAMPLE VOLUME in mL SS	0.100	PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR DF	1.000	PU 239 AEA FRAC (C239)	0.000
27632	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	1605
Test Code	DIGEST DILUTION FACTOR DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cp (Bkg)	0.100
Matrix	DETECTOR NUMBER	13	PU 236 cpm	35.880
LIQUID	EFFICIENCY FACTOR EFF	0.3166	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	0.000
98006245	TRACER PREPARATION VALUE (dpm/mL)	2270.00	AEA COUNT TIME	480
Retun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1804.08	Pu 239/240 $\mu\text{Ci/L}$ =	< 2.932E-02
0	PU-238 TRACER VALUE (dpm/m)	0.00		

Sample Prep:
N/ASample #
WL27632

Decay Time = Date Counted - Tracer Preparation Date

Instrument Code
WB27810

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(-ln2 * Decay Time/1040.95)}]

Prepared By
SZC

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

Chemist
JFRPu 239/240 $\mu\text{Ci/L}$ = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) / (DF) / (DDF) / [(C236) (SS) (D g/L) (2220000 dpm/ μCi)]Analyst
RRO

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

Date Complete
12/31/98Pu 238 $\mu\text{Ci/L}$ = [(Pu 238 dpm) / (DF) / (DDF) / (1000 mL/L)] / [(Pu-238 Tracer Recovery / 100) (2220000 dpm/ μCi) (D g/L) (SS)]Analysis Date
12/30/98

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

Analysis Time
11:15 AMPu 239/240 $\mu\text{Ci/mL}$ < 2.93E-05

Relative Counting Error = 100.0%

DETECTION
LEVELS
in $\mu\text{Ci/mL}$ Sample Point
BY-105 (2)

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 238 $\mu\text{Ci/mL}$ < 2.93E-05

Relative Counting Error = 100.0%

Pu 236 Tracer Recovery = 88.2%

Pu 239/240
2.93E-05
Pu 238
2.93E-05

Analyst	RRO	Date:	31-Dec-98
Signature of Chemist: <i>John Relyea</i>	JFR	Date:	4 Jan 99

BLANK.WB1 REV 1.0

643128ML

WORKBOOK PAGE: SAM3

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****SAMPLE**

TYPE	DATE COUNTED	DEC-30-98	PU 238 AEA FRAC (C236)	0.828
SAMPLE	SAMPLE VOLUME in mL	SS	0.100 PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	1.000 PU 239 AEA FRAC (C239)	0.092
27632	TRACER VOLUME in mL	SPKV	0.100 TOTAL AT COUNTS	472
0.01000	DIGEST DILUTION FACTOR	DDF	1.0000 AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.100
Matrix	DETECTOR NUMBER	13	PU 236 cpm	9.130
LIQUID	EFFICIENCY FACTOR	EFF	0.317 PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	1.020
98006245	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Rerun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1804.081	Pu 239/240 μ Ci/L	< 1.2992E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample

Decay Time = Date Counted - Tracer Preparation Date

S98T003384

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times$ [e to the power of {(ln2 * Decay Time)/(1040.95)}]

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WB27810

Pu 239/240 μ Ci/L = (C239)/(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(D g/L)(2220000 dpm/ μ Ci)]

Prepared By

Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

SZC

Pu 238 μ Ci/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/ μ Ci)(D g/L)(SS)]

Chemist

Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analysis

RRO

Date Complete

12/31/98

Pu 239/240 μ Ci/mL < 1.30E-04

DETECTION

Analysis Date

Relative Counting Error = 9.3%

LEVELS

12/30/98

in μ Ci/mL

Analysis Time

NOTE: Pu 238 Result is a LESS THAN Value.

Pu 239/240

11:15 AM

Pu 238 μ Ci/mL < 1.30E-04

1.30E-04

Sample Point

Relative Counting Error = 100.0%

Pu 238

BY-105 (2)

Pu 236 Tracer Recovery = 22.7%

1.30E-04

Analyst:	RRO	Date:	31-Dec-98
Signature of Chemist:	JFR	Date:	4 Jan 99

SAMPLE.WB1 REV 1.0

943128ML

WORKBOOK PAGE: DUP4

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**LIQUID /****DUP**

DATE COUNTED	DEC-30-98	PU 236 AEA FRAC (C236)	0.627	
DUP	SAMPLE VOLUME in mL SS	0.100	PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR DF	1.000	PU 239 AEA FRAC (C239)	0.267
27632	TRACER VOLUME in mL SPKV	0.100	TOTAL AT COUNTS	502
Test Code	DIGEST DILUTION FACTOR DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)	0.100
Matrix	DETECTOR NUMBER	13	PU 236 cpm	7.450
LIQUID	EFFICIENCY FACTOR EFF	0.317	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm	3.170
98006245	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME	480
Remun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1804.081	Pu 239/240 μ Ci/L	3.4606E-01
0	PU-238 TRACER VALUE (dpm/mL)	0.000		

Sample Prep

N/A

Sample

Decay Time = Date Counted - Tracer Preparation Date

S98T003384

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of {(ln2 * Decay Time/1040.95)}]

Instrument Code

Pu 236 Tracer Recovery = (Total AT counts / TC - Bkg) * C236 * 100 / (Pu-236 Decay Corr'd Value * SPKV * EFF)

WB27810

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) * (SPKV) / (1000 mL/L) * (DF) / (DDF) / [(C236) * (D g/L) / (2220000 dpm/ μ Ci)]

Prepared By

Pu 238 dpm = [(Total AT counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)

SZC

Pu 238 μ Ci/L = [(Pu 238 dpm) / (DF) * (DDF) / (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) * (2220000 dpm/ μ Ci) / (D g/L) / (SS)]

Chemist

Relative Counting Error = Square Root of [(1 / (Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100

JFR

Analyst

RRO

Date Complete

12/31/98

Pu 239/240 μ Ci/mL

Analysis Date

12/30/98

Relative Counting Error = 6.0%

Analysis Date

12/30/98

NOTE: Pu 238 Result is a LESS THAN Value.

Analysis Time

11:15 AM

Pu 238 μ Ci/mL

Sample Point

BY-105 (2)

< 2.13E-04

Relative Counting Error = 100.0%

Pu 236 Tracer Recovery = 18.3%

DETECTION

LEVELS

in μ Ci/mL

Pu 239/240

2.13E-04

Pu 238

2.13E-04

Analyst:	RRO	Date:	31-Dec-98
Signature of Chemist:	JFR	Date:	9 Jan 99
SAMPLE WB1 REV 1.0	943128ML		

5017

222-S Analytical Laboratory
 GENERAL ALPHA ENERGY ANALYSIS
 Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
 WL27632-STD(PU)
 File ID: 1a1883.CNF

Counted on: 12/30/98 @18:12
 Detector: AEA1
 Geometry number: 1
 Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1291.4	1291.4	361.760	361.760	10.000	2.850	5.000	1.212
2	44.3	44.3	301.433	301.191	10.000	2.763	5.000	1.598
3	76.9	76.9	287.729	287.592	8.000	2.153	4.000	1.407
4?	18.2	18.2	270.708	270.184	10.000	2.362	5.000	0.967
5	2332.3	2332.3	229.030	229.030	10.000	2.971	5.000	1.736

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff. FWHM				
1	Pu236	0.372	5.755	5.754	0.0010.01	35.37	1.5	149.4	0.673E-04
2	Pu238	0.011	5.487	5.476	0.0110.01	1.02	9.2	5.8	0.263E-05
	Am241		5.479	5.476	0.003			4.5	0.202E-05
3	Th228	0.016	5.400	5.413	-0.0130.01	1.56	8.0	9.1	0.409E-05
4	????			5.333		0.49	16.2		
5	Pu239	0.571	5.147	5.144	0.0030.01	54.23	1.2	224.4	0.101E-03
	Pu240		5.144	5.144	0.000			224.4	0.101E-03

Totals:			0.970	<--valid peaks only-->		92.17			

DETECTOR CALIBRATION

Energy(MEV) = 4.090 + (0.0046)*Channel
 Energy range (MeV): 4.090 TO 6.446
 Efficiency = 0.2416 CPM/DPM
 (Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	45597.0	100.000
Smoothed	45597.0	100.000
Composite fit	44481.9	97.554
Residuals	1115.1	2.446

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 15145.6

.
.5
.....5.
.....5
.....5
.....5
.....5..

4
4
4
4
3.
3
2
2

1
...1
.....1.
.....1
.....1
..

Raw Data Dump for AEA Spectrum: 1a1883.CNF

HNF-1663 REV. 0

1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	1.	0.	0.	0.	0.	0.	0.	1.	2.	0.
31	0.	1.	1.	0.	0.	0.	0.	0.	0.	0.
41	0.	2.	1.	0.	0.	0.	0.	0.	0.	0.
51	1.	2.	1.	0.	1.	1.	0.	0.	0.	0.
61	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
71	0.	1.	0.	0.	0.	0.	2.	0.	2.	0.
81	1.	1.	0.	1.	0.	2.	0.	0.	0.	1.
91	2.	2.	1.	0.	0.	2.	0.	0.	0.	2.
101	0.	3.	0.	1.	1.	1.	2.	1.	3.	1.
111	0.	1.	2.	2.	0.	0.	0.	1.	0.	2.
121	0.	1.	1.	3.	0.	1.	1.	0.	1.	1.
131	0.	1.	0.	1.	3.	1.	1.	0.	1.	0.
141	3.	2.	1.	1.	2.	1.	1.	1.	3.	2.
151	5.	3.	4.	1.	3.	1.	4.	4.	1.	3.
161	2.	3.	2.	2.	4.	3.	3.	1.	4.	1.
171	7.	8.	6.	9.	5.	6.	11.	12.	8.	15.
181	10.	9.	9.	15.	12.	16.	17.	15.	25.	11.
191	27.	20.	38.	25.	35.	29.	39.	31.	37.	49.
201	51.	53.	67.	74.	87.	101.	106.	109.	140.	151.
211	184.	193.	213.	278.	309.	418.	462.	587.	674.	714.
221	752.	837.	916.	1167.	1483.	1816.	2220.	2528.	2832.	2783.
231	2205.	1379.	502.	132.	18.	6.	0.	0.	3.	3.
241	1.	3.	1.	3.	3.	1.	0.	4.	3.	5.
251	3.	9.	3.	6.	3.	3.	5.	9.	11.	5.
261	4.	5.	6.	10.	11.	13.	15.	16.	18.	33.
271	23.	22.	19.	11.	15.	10.	19.	15.	18.	25.
281	23.	37.	46.	59.	61.	79.	83.	108.	102.	60.
291	25.	18.	12.	13.	15.	16.	22.	20.	26.	34.
301	27.	25.	31.	25.	26.	17.	12.	12.	8.	4.
311	7.	7.	7.	7.	9.	7.	6.	5.	10.	11.
321	3.	10.	19.	19.	14.	9.	19.	33.	17.	16.
331	28.	36.	33.	35.	45.	33.	50.	62.	71.	74.
341	89.	110.	133.	150.	188.	193.	238.	283.	327.	455.
351	596.	654.	682.	667.	623.	575.	687.	844.	952.	1184.
361	1462.	1595.	1495.	1095.	579.	169.	40.	5.	0.	0.
371	0.	0.	0.	1.	0.	0.	0.	0.	0.	1.
381	0.	0.	0.	0.	1.	0.	0.	1.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
411	0.	0.	1.	1.	0.	0.	1.	1.	0.	0.
421	0.	1.	3.	1.	5.	0.	2.	3.	1.	0.
431	1.	4.	2.	0.	0.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
451	0.	0.	1.	0.	0.	1.	0.	2.	1.	1.
461	1.	3.	0.	2.	1.	1.	1.	2.	5.	4.
471	3.	6.	10.	9.	13.	8.	7.	7.	6.	4.
481	0.	1.	0.	1.	0.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL27632-BLNK(PU)
File ID: 2a2080.CNF

Counted on: 12/30/98 @18:13
Detector: AEA2
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1367.2	1367.2	362.508	362.508	10.000	3.116	5.000	1.473
2?	20.6	20.6	304.050	303.982	10.000	0.866	5.000	0.493
3	77.2	77.2	288.225	288.183	10.000	2.802	5.000	1.983
4?	15.2	15.2	270.357	269.926	12.000	2.476	6.000	1.650

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err @95	d/m	Activity uCi/ea
		Exp.	Obs.	Diff.	c/m			
1	Pu236	0.943	5.755	5.744	0.0110.01	35.88	1.5	154.6
2		????		5.475		0.38	15.1	
3	Th228	0.043	5.400	5.402	-.0020.01	1.62	7.3	9.6
4		????		5.318		0.32	20.9	
Totals:		0.985	<--valid peaks only-->			37.50		

DETECTOR CALIBRATION

Energy(MEV) = 4.076 + (0.0046)*Channel
Energy range (MeV): 4.076 TO 6.431
Efficiency = 0.2368 CPM/CPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	18270.0	100.000
Smoothed	18269.3	99.996
Composite fit	18336.5	100.364
Residuals	-66.5	-0.364

Analyzed by: _____

MB

Spectrum 2a2080.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 7818.0

4
4
4
..3
23
2
2

1
.. 1
.....1
.....1
.....1
.....1
..1...

Raw Data Dump for AEA Spectrum: 2a2080.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
31	0.	0.	0.	0.	1.	0.	1.	0.	0.	0.
41	0.	1.	1.	0.	0.	0.	0.	0.	0.	0.
51	1.	0.	0.	0.	0.	0.	0.	1.	1.	0.
61	0.	1.	0.	0.	0.	0.	1.	0.	0.	2.
71	0.	0.	0.	0.	0.	0.	3.	0.	0.	0.
81	0.	0.	1.	0.	0.	0.	0.	0.	0.	1.
91	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
101	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
111	0.	0.	0.	1.	1.	0.	0.	0.	0.	0.
121	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
131	0.	0.	0.	0.	0.	0.	0.	2.	0.	0.
141	0.	0.	1.	0.	0.	0.	1.	0.	0.	0.
151	0.	0.	1.	0.	1.	1.	1.	1.	0.	1.
161	0.	0.	0.	0.	1.	0.	0.	0.	0.	1.
171	0.	0.	0.	0.	1.	0.	0.	0.	1.	0.
181	0.	2.	0.	0.	0.	0.	0.	0.	0.	1.
191	0.	1.	0.	0.	0.	0.	1.	0.	0.	1.
201	1.	1.	2.	1.	3.	0.	0.	0.	2.	1.
211	3.	3.	4.	5.	3.	5.	1.	4.	3.	1.
221	4.	5.	5.	2.	2.	8.	8.	5.	9.	7.
231	6.	5.	4.	3.	2.	1.	0.	2.	1.	1.
241	1.	1.	2.	2.	1.	2.	1.	0.	0.	2.
251	2.	2.	5.	1.	2.	3.	0.	4.	1.	3.
261	7.	4.	5.	7.	10.	4.	14.	11.	24.	22.
271	25.	13.	11.	17.	12.	4.	6.	10.	10.	25.
281	26.	33.	34.	48.	73.	58.	88.	96.	99.	76.
291	50.	23.	6.	5.	5.	9.	5.	7.	6.	7.
301	7.	5.	18.	19.	15.	13.	5.	4.	0.	1.
311	0.	1.	2.	1.	1.	1.	1.	1.	3.	6.
321	8.	3.	5.	3.	3.	3.	8.	8.	12.	11.
331	7.	15.	9.	9.	17.	18.	17.	27.	19.	25.
341	49.	63.	51.	81.	122.	138.	184.	199.	289.	351.
351	437.	563.	616.	654.	656.	671.	663.	812.	860.	1154.
361	1393.	1557.	1689.	1496.	1073.	525.	211.	56.	3.	5.
371	3.	1.	0.	0.	0.	1.	0.	2.	0.	0.
381	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
391	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
401	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	1.	1.	0.	1.
421	0.	1.	0.	1.	0.	0.	1.	0.	0.	0.
431	0.	0.	0.	2.	1.	0.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	1.	1.	0.	0.	0.	0.	0.	1.
461	0.	1.	1.	2.	2.	0.	2.	6.	1.	4.
471	5.	2.	11.	8.	12.	7.	10.	12.	4.	2.
481	2.	4.	1.	0.	0.	0.	0.	1.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3384-SAM(PU)
File ID: 3a3372.CNF

Counted on: 12/30/98 @18:13
Detector: AEA3
Geometry number: 1
Count time: 28804. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	334.3	334.3	362.532	362.532	10.000	3.342	5.000	1.513
2	29.9	29.9	287.874	287.872	8.000	3.945	4.000	2.772
3	37.6	37.6	228.865	228.865	14.000	4.532	7.000	2.429

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			FWHM	Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.					
1	Pu236	0.828	5.755	5.757	-0.0020.02	9.13	3.0	40.9	0.184E-04	
2	Th228	0.063	5.400	5.414	-0.0140.02	0.70	10.7	4.3	0.195E-05	
3	Pu239	0.092	5.147	5.142	0.0050.02	1.02	8.9	4.5	0.202E-05	
	Pu240		5.144	5.142	0.002			4.5	0.202E-05	

Totals:		0.984	<--valid peaks only-->				10.85			

DETECTOR CALIBRATION

Energy (MeV) = 4.089 + (0.0046)*Channel
Energy range (MeV): 4.089 TO 6.445
Efficiency = 0.2276 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	5293.0	100.000
Smoothed	5293.0	100.000
Composite fit	5208.3	98.400
Residuals	84.7	1.600

Analyzed by: _____

Spectrum 3a3372.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 1951.4

3
 ..3.
3
 .3

.
 2
2
 ...2
 .

1
 .. 1
1
 .....1.
1
 ...1...

Raw Data Dump for AEA Spectrum: 3a3372.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
21	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
31	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
41	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
51	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
61	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
71	0.	0.	0.	0.	0.	1.	0.	0.	0.	0.
81	1.	1.	0.	0.	0.	0.	0.	0.	0.	1.
91	0.	0.	1.	0.	0.	1.	0.	1.	0.	0.
101	0.	0.	0.	0.	0.	1.	0.	0.	1.	0.
111	0.	0.	1.	0.	0.	0.	0.	0.	1.	1.
121	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
131	0.	1.	1.	1.	1.	0.	0.	0.	0.	0.
141	0.	1.	0.	0.	0.	0.	0.	0.	0.	0.
151	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
161	0.	0.	0.	0.	2.	0.	1.	0.	0.	0.
171	0.	0.	0.	0.	0.	1.	0.	2.	0.	1.
181	0.	1.	0.	1.	1.	0.	1.	0.	2.	0.
191	1.	0.	1.	0.	0.	0.	0.	0.	1.	0.
201	2.	1.	3.	2.	0.	2.	3.	1.	0.	1.
211	4.	3.	2.	5.	6.	6.	10.	9.	13.	13.
221	16.	16.	17.	38.	29.	28.	39.	47.	41.	39.
231	38.	30.	18.	5.	2.	2.	1.	0.	0.	0.
241	0.	0.	1.	2.	1.	4.	0.	2.	1.	3.
251	0.	1.	2.	1.	0.	3.	1.	2.	0.	1.
261	1.	1.	5.	4.	2.	1.	5.	3.	9.	11.
271	10.	3.	3.	5.	2.	6.	7.	4.	5.	8.
281	8.	11.	13.	17.	26.	27.	37.	31.	47.	16.
291	20.	6.	11.	2.	5.	3.	4.	3.	7.	6.
301	8.	5.	5.	2.	6.	2.	3.	3.	2.	2.
311	0.	1.	1.	1.	0.	1.	1.	0.	1.	0.
321	0.	4.	0.	2.	0.	3.	0.	1.	5.	3.
331	4.	2.	3.	2.	5.	2.	8.	3.	6.	10.
341	4.	13.	20.	15.	20.	45.	49.	55.	78.	87.
351	115.	123.	183.	166.	177.	166.	177.	186.	226.	277.
361	344.	379.	412.	358.	267.	160.	66.	19.	2.	0.
371	0.	0.	0.	0.	1.	1.	0.	0.	1.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	1.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	1.	2.	0.
421	0.	1.	0.	1.	0.	0.	0.	1.	1.	0.
431	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
461	0.	0.	1.	0.	1.	0.	0.	1.	0.	1.
471	1.	2.	1.	2.	2.	4.	3.	5.	4.	1.
481	1.	1.	1.	0.	0.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.								

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T3384-DUP(PU)
File ID: 4a4369.CNF

Counted on: 12/30/98 @18:14
Detector: AEA4
Geometry number: 1
Count time: 28801. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	288.3	288.3	359.858	359.858	10.000	2.887	5.000	1.364
2	23.7	23.7	285.659	285.657	10.000	3.569	5.000	2.026
3	98.8	98.8	226.529	226.529	12.000	4.437	6.000	1.687

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err c/m @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.				
1	Pu236	0.627	5.755	5.744	0.0110.01	7.45	3.3	33.8	0.152E-04
2	Th228	0.049	5.400	5.403	-0.0030.02	0.59	11.7	3.7	0.166E-05
3	Pu239	0.267	5.147	5.131	0.0160.02	3.17	5.0	14.1	0.635E-05
	Pu240		5.144	5.131	0.013			14.1	0.635E-05

Totals:		0.943	<--valid peaks only-->			11.20			

DETECTOR CALIBRATION

Energy (MEV) = 4.089 + (0.0046)*Channel
Energy range (MeV): 4.089 TO 6.444
Efficiency = 0.2245 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	5700.0	100.000
Smoothed	5700.0	100.000
Composite fit	5376.5	94.324
Residuals	323.5	5.676

Analyzed by: _____

MB

Spectrum 4a4369.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 1600.7

.
.
3.
..3.
.....3
.....3
.....3
..3.

.
.2
.....2
2.
.

1
. 1
..... 1
.....1...
..... 1
.....1.
..

Raw Data Dump for AEA Spectrum: 4a4369.CNF										
1	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	2.	0.
21	0.	0.	0.	1.	0.	0.	1.	0.	0.	0.
31	0.	1.	0.	0.	0.	0.	0.	1.	0.	1.
41	0.	0.	0.	0.	1.	0.	0.	0.	0.	0.
51	0.	0.	0.	1.	0.	0.	0.	0.	0.	0.
61	1.	0.	1.	2.	0.	2.	1.	0.	0.	0.
71	2.	1.	0.	1.	0.	1.	1.	0.	0.	2.
81	0.	3.	0.	0.	0.	0.	1.	1.	1.	1.
91	2.	0.	2.	1.	0.	0.	2.	2.	1.	2.
101	1.	2.	2.	0.	1.	0.	1.	1.	2.	0.
111	2.	2.	0.	0.	2.	1.	1.	3.	1.	2.
121	1.	2.	0.	0.	1.	0.	0.	1.	0.	2.
131	3.	2.	2.	3.	2.	1.	0.	1.	1.	3.
141	2.	0.	1.	1.	2.	0.	2.	1.	2.	4.
151	3.	3.	1.	4.	0.	2.	5.	3.	1.	3.
161	3.	0.	3.	3.	2.	3.	1.	4.	4.	0.
171	2.	1.	5.	1.	3.	2.	5.	6.	4.	6.
181	2.	3.	6.	9.	3.	3.	6.	2.	8.	5.
191	6.	5.	7.	9.	7.	4.	6.	6.	11.	9.
201	12.	11.	12.	9.	10.	17.	16.	17.	10.	22.
211	19.	22.	35.	27.	36.	37.	45.	53.	55.	61.
221	55.	76.	63.	98.	102.	113.	110.	112.	81.	62.
231	42.	15.	5.	2.	1.	3.	1.	0.	0.	2.
241	1.	1.	0.	2.	1.	4.	0.	2.	2.	2.
251	2.	4.	1.	1.	2.	2.	1.	2.	2.	1.
261	3.	2.	3.	1.	7.	3.	7.	8.	8.	9.
271	5.	1.	2.	4.	1.	8.	6.	8.	3.	11.
281	7.	19.	21.	23.	24.	34.	21.	21.	11.	2.
291	3.	3.	4.	4.	3.	5.	3.	11.	3.	4.
301	7.	7.	3.	3.	1.	2.	0.	0.	2.	2.
311	1.	0.	0.	0.	1.	1.	0.	1.	0.	2.
321	0.	2.	0.	0.	0.	1.	1.	4.	2.	6.
331	1.	3.	3.	3.	9.	9.	7.	6.	10.	11.
341	6.	23.	28.	34.	46.	57.	68.	85.	108.	151.
351	127.	151.	119.	124.	159.	188.	208.	272.	312.	356.
361	336.	261.	134.	47.	14.	2.	0.	1.	0.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	1.	0.	1.
421	0.	0.	0.	1.	0.	1.	0.	0.	1.	0.
431	0.	0.	0.	0.	0.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	0.	0.	0.	0.	1.	0.
461	0.	0.	1.	0.	1.	0.	3.	2.	1.	3.
471	5.	2.	0.	3.	2.	3.	1.	1.	0.	0.
481	1.	0.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	1.	0.	0.	0.
511	0.	0.								

LABCORE Completed Worklist Report for Worklist# 28059

Analyst: rro Instrument: AB13 Book#: _____

Method: LA-953-104 Rev/Mod _____

Worklist Comment: BY105 FIELD BLANK FOR @PU23901 RTS

Seq Type	Sample#	R A	Test	Matrix	Actual	Found	DL or Yield	Unit
1 STD	0		@PU23901 PU23901 LIQUID		1.12E-04	1.27E-4	113.353	% Recovery
1 STD	0		@PU23901 PU23901T LIQUID		100	6.37E+01	83.700	% Recovery
1 STD	0		@PU23901 PU23901E LIQUID		1.00	1.97E+00	1.970	% Ct Error
2 BLNK	0		@PU23901 PU23901 LIQUID		1	<6.56E-7		uCi/mL
2 BLNK	0		@PU23901 PU23901T LIQUID		100	7.89E+01	78.900	% Recovery
2 BLNK	0		@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000	uCi/mL
3 SAMPLE	S98T002391	0	@PU23901 PU23901 LIQUID	N/A	<	6.71E-07	6.71e-007	uCi/mL
3 SAMPLE	S98T002391	0	@PU23901 PU23901T LIQUID	N/A		7.88E+01		% Recovery
3 SAMPLE	S98T002391	0	@PU23901 PU23901E LIQUID	N/A		1.00E+02		% Ct Error
4 DUP	S98T002391	0	@PU23901 PU23901 LIQUID		<6.71E-7	<5.35E-7		RPD
4 DUP	S98T002391	0	@PU23901 PU23901T LIQUID		100	9.23E+01	92.300	% Recovery
4 DUP	S98T002391	0	@PU23901 PU23901E LIQUID		1.00	1.00E+02	100.000	% Ct Error

Final page for worklist# 28059

Analyst Signature _____ Date _____

Analyst Signature _____ Date _____

37 Feb 99
Reviewer Signature _____ Date _____

Units shown for QC (BLK/BKG) may not reflect the actual units.

LABCORE Data Entry Template for Worklist# 28059

Analyst: RRO Instrument: PU01 13 Book# 28057

Method: LA-953-104 Rev/Mod B-1

Worklist Comment: BY105 FIELD BLANK FOR @PU23901 RTS

S Type	Sample#	R A	Test	Matrix	Group#	Project
1 STD			@PU23901	LIQUID		
2 BLNK			@PU23901	LIQUID		
3 SAMPLE	S98T002391 0		@PU23901	LIQUID	98000436	BY-105 (2)
	Analytes Requested: PU23901 , PU23901E, PU23901T					
4 DUP	S98T002391 0		@PU23901	LIQUID		

Final page for worklist # 28059

RRO'Veal 2-2-99
Analyst Signature Date

KRB 2/3/99
Analyst Signature Date
Plus 2 Cham 2/5/99

Data Entry Comments:

S = Worklist Slot Number, R = Replicate Number, A = Aliquot Code.

WORKBOOK PAGE: STD1

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID**

				STD
TYPE	DATE COUNTED	FEB-02-99		Pu 238 AEA FRAC (C238)
STD	SAMPLE VOLUME in mL	SS	1.000	Pu 238 AEA FRAC (C238)
Work List	SAMPLE DILUTION FACTOR	DF	1.000	Pu 239 AEA FRAC (C239)
28059	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS
Test Code	DIGEST DILUTION FACTOR	DDF	1.000	AT COUNT TIME (MIN)
@PU23901	TRACER BOOK NO	145B43		BACKGROUND in cpm (Bkg)
Matrix	DETECTOR NUMBER	13		Pu 236 cpm
LIQUID	EFFICIENCY FACTOR	EFF	0.317	Pu 238 cpm
Batch Number	TRACER PREPARATION DATE	01/19/98		Pu 239 cpm
99000331	TRACER PREPARATION VALUE (dpm/mL)	2270.000		AEA COUNT TIME
Refund	PU-238 DECAY CORR'D VALUE (dpm/mL)	1763.696		Pu 239/240 µCi/L
0	PU-238 TRACER VALUE (dpm/mL)	0.000		
Sample Prep	STANDARD BOOK NO	28B57		
N/A	STANDARD VALUE in µCi/mL	1.120E-04		
Sample #				
WL28059-STD				
Instrument Code	Decay Time = Date Counted - Tracer Preparation Date			
WB27810	Pu-236 Decay Corr'd Value = Pu-236 Preparation Value * [e to the power of ((-ln2 * Decay Time/1040.95))]			
Prepared By	Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) * 1/EFF * C236 * 100 / Pu-236 Decay Corr'd Value * SPKV			
MCB	Pu 239/240 µCi/L = (C239)(Pu 236 Decay Corr'd Value)(SPKV)(1000mL/L)(DF)(DDF) / [(C236)(SS)(2220000 dpm/µCi)]			
Chemist	Pu 238 dpm = [(Total AT Counts / TC) - Bkg * 1/EFF * C238] - (Pu-238 Tracer Value * SPKV * Pu 236 Tracer Recovery / 100)			
JFR	Pu 238 µCi/L = [(Pu 238 dpm)(DF)(DDF)(1000mL/L)] / [(Pu-236 Tracer Recovery / 100)(2220000 dpm/µCi)(D g/L)(SS)]			
Analyst	Relative Counting Error = Square Root of [(1/(Pu 236 cpm * min)) + (1 / (Pu 238 or 239/240 cpm * min))] * 1.96 * 100			
RRO				
Data Complete				
02/03/99	Pu 239/240 µCi/mL	1.27E-04	DETECTION LEVELS in µCi/mL Pu 239/240 8.05E-06	
Analysis Date	Relative Counting Error	= 2.0%		
02/02/99				
Analysis Time				
03:05 PM				
Sample Print				
BY-105	Pu 236 Tracer Recovery	= 83.7%		

Analyst:	RRO	Date:	03-Feb-99
Signature of Chemist:	JFR	Date:	3 Feb 99

STANDARD.WB1 REV 1.0

LA-943-128ML

5032

WORKBOOK PAGE: BLANK2

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))

L

BLNK

BLNK	SAMPLE VOLUME in mL	SS	5.000	PU 238 AEA FRAC (C238)	0.921
28059	SAMPLE DILUTION FACTOR	DF	1.000	PU 238 AEA FRAC (C238)	0.000
	TRACER VOLUME in mL	SPKV	0.100	PU 239 AEA FRAC (C239)	0.000
	DIGEST DILUTION FACTOR	DDF	1.0000	TOTAL AT COUNTS	1437
@PU23901	TRACER BOOK NO		145B43	AT COUNT TIME (MIN)	30
	DETECTOR NUMBER		13	BACKGROUND in cp (Bkg)	0.070
LIQUID	EFFICIENCY FACTOR	EFF	0.3166	PU 238 cpm	32.920
Batch Number	TRACER PREPARATION DATE		01/19/98	PU 238 cpm	0.000
99000331	TRACER PREPARATION VALUE (dpm/mL)		2270.00	PU 239 cpm	0.000
	PU-236 DECAY CORR'D VALUE (dpm/mL)		1763.70	AEA COUNT TIME	480
0	PU-238 TRACER VALUE (dpm/mL)		0.00	Pu 239/240 $\mu\text{Ci/L}$ =	< 6.560E-04

Sample Prep

N/A

Sample Prep

Decay Time = Date Counted - Tracer Preparation Date

WL28059-BLNK

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times [e \text{ to the power of } \{(-\ln 2 \times \text{Decay Time}/1040.95)\}]$

Instrument Code

Pu 236 Tracer Recovery = $(\text{Total AT Counts} / \text{TC} - \text{Bkg}) \times \text{C236} \times 100 / (\text{Pu-236 Decay Corr'd Value} \times \text{SPKV} \times \text{EFF})$

WB27810

Pu 239/240 $\mu\text{Ci/L}$ = $(\text{C239} / (\text{Pu 236 Decay Corr'd Value} \times (\text{SPKV} / (1000 \text{ mL/L}) \times (\text{DF}) \times (\text{DDF}) / ((\text{C236}) \times (\text{D g/L}) \times (2220000 \text{ dpm}/\mu\text{Ci})))$

Prepared By

Pu 238 dpm = $[(\text{Total AT Counts} / \text{TC}) - \text{Bkg} \times 1/\text{EFF} \times \text{C238}] - (\text{Pu-238 Tracer Value} \times \text{SPKV} \times \text{Pu 236 Tracer Recovery} / 100)$

MCB

Pu 238 $\mu\text{Ci/L}$ = $[(\text{Pu 238 dpm}) \times (\text{DF}) \times (\text{DDF}) \times (1000 \text{ mL/L})] / [(\text{Pu-236 Tracer Recovery} / 100) \times (2220000 \text{ dpm}/\mu\text{Ci}) \times (\text{D g/L}) \times (\text{SS})]$

Chemist

Relative Counting Error = Square Root of $\{[1 / (\text{Pu 238 cpm} \times \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

JFR

Analyst

RRO

Date completed

02/03/99

Analysis Date

02/02/99

Analysis Time

03:05 PM

Sample Point

BY-105

Pu 239/240 $\mu\text{Ci/mL}$	< 6.56E-07	DETECTION LEVELS
Relative Counting Error =	100.0%	in $\mu\text{Ci/mL}$
NOTE: Pu 238 Result is a LESS THAN Value.		Pu 239/240
Pu 238 $\mu\text{Ci/mL}$	< 6.56E-07	6.56E-07
Relative Counting Error =	100.0%	Pu 238
Pu 236 Tracer Recovery =	78.9%	6.56E-07

Analyst:	RRO	Date:	03-Feb-99
Signature of Chemist:	JFR	Date:	3 Feb 99

BLANK.WB1 REV 1.0

943128ML

5033

WORKBOOK PAGE: SAM3

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)**LIQUID /**

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0)				LIQUID /	SAMPLE
DATE COUNTED		FEB-02-99	PU 236 AEA FRAC (C236)		0.901
SAMPLE	SAMPLE VOLUME in mL	SS	5.000	PU 238 AEA FRAC (C238)	0.016
Work List	SAMPLE DILUTION FACTOR	DF	1.000	PU 239 AEA FRAC (C239)	0.000
28059	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1468
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.070
Matrix	DETECTOR NUMBER	13	PU 236 cpm		32.470
LIQUID	EFFICIENCY FACTOR	EFF	0.317	PU 238 cpm	0.590
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.000
99000331	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
ReRun	PU-236 DECAY CORR'D VALUE (dpm/mL)	1763.696	Pu 239/240 μ Ci/L = < 6.7100E-04		
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample

Decay Time = Date Counted - Tracer Preparation Date

S98T002391

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times [e^{-(\ln 2 \times \text{Decay Time}/1040.95)}]$

Instrument Code

Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times$ C236 $\times$ 100 / (Pu-236 Decay Corr'd Value $\times$ SPKV $\times$ EFF)

WB27810

Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) $\times$ (SPKV) / (1000 mL/L) / (DF) / (DDF) / ((C236) / (SS) / (D g/L) / (2220000 dpm/ μ Ci))

Prepared By

Pu 238 dpm = ((Total AT Counts / TC) - Bkg $\times$ 1/EFF $\times$ C238) - (Pu-238 Tracer Value $\times$ SPKV $\times$ Pu 236 Tracer Recovery / 100)

MCB

Pu 238 μ Ci/L = ((Pu 238 dpm) / (DF) / (DDF) / (1000 mL/L)) / ((Pu-236 Tracer Recovery / 100) / (2220000 dpm/ μ Ci) / (D g/L) / (SS))

Chemist

Relative Counting Error = Square Root of $\{[1 / (\text{Pu 236 cpm} \times \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

JFR

Analyst

RRO

Data Complete

02/03/99

Pu 239/240 μ Ci/mL

< 6.71E-07

Analysis Date

Relative Counting Error

= 100.0%

02/02/99

Analysis Time

NOTE: Pu 238 Result is a LESS THAN Value.

03:05 PM

Pu 238 μ Ci/mL

< 6.71E-07

Sample Point

Relative Counting Error

= 11.8%

BY-105

Pu 236 Tracer Recovery

= 78.8%

DETECTION

LEVELS

in μ Ci/mL

Pu 239/240

6.71E-07

Pu 238

6.71E-07

Analyst:	RRO	Date:	03-Feb-99
Signature of Chemist:	JFR	Date:	3 Feb 99

SAMPLE WB1 REV 1.0

John Relyea
943128ML

5034

WORKBOOK PAGE: DUP4

Pu 238 and 239/240 : LA-943-128 (VOID) or LA-953-104(B-0))**LIQUID /****DUP**

TYPE	DATE COUNTED	FEB-02-99	PU 236 AEA FRAC (C236)	0.965	
DUP	SAMPLE VOLUME in mL	SS	5.000	PU 238 AEA FRAC (C238)	0.000
Work List	SAMPLE DILUTION FACTOR	DF	1.000	PU 239 AEA FRAC (C239)	0.000
28059	TRACER VOLUME in mL	SPKV	0.100	TOTAL AT COUNTS	1604
Test Code	DIGEST DILUTION FACTOR	DDF	1.0000	AT COUNT TIME (MIN)	30
@PU23901	TRACER BOOK NO	145B43	BACKGROUND in cpm (Bkg)		0.070
Main	DETECTOR NUMBER	13	PU 236 cpm		25.930
LIQUID	EFFICIENCY FACTOR	EFF	0.317	PU 238 cpm	0.000
Batch Number	TRACER PREPARATION DATE	01/19/98	PU 239 cpm		0.000
99000331	TRACER PREPARATION VALUE (dpm/mL)	2270.000	AEA COUNT TIME		480
Return	PU-236 DECAY CORR'D VALUE (dpm/mL)	1763.696	Pu 239/240 μ Ci/L =	<	5.3529E-04
0	PU-238 TRACER VALUE (dpm/mL)	0.000			

Sample Prep

N/A

Sample #

S98T002391

Instrument Code

WB27810

Prepared By

MCB

Chemist

JFR

Analyst

RRO

Date Complete

02/03/99

Analysis Date

02/02/99

Analysis Time

03:05 PM

Sample Point

BY-105

Decay Time = Date Counted - Tracer Preparation Date

Pu-236 Decay Corr'd Value = Pu-236 Preparation Value $\times$ [e to the power of $\{(-\ln 2 \times \text{Decay Time}) / 1040.95\}$]Pu 236 Tracer Recovery = (Total AT Counts / TC - Bkg) $\times$ C236 $\times$ 100 / (Pu-236 Decay Corr'd Value $\times$ SPKV $\times$ EFF)Pu 239/240 μ Ci/L = (C239) / (Pu 236 Decay Corr'd Value) $\times$ (SPKV $\times$ (1000 mL/L) $\times$ (DF) $\times$ (DDF) / [(C236) $\times$ (SS) $\times$ (D g/L) $\times$ (2220000 dpm/ μ Ci)])Pu 238 dpm = [(Total AT Counts / TC) - Bkg $\times$ 1/EFF $\times$ C238] - (Pu-238 Tracer Value $\times$ SPKV $\times$ Pu 236 Tracer Recovery / 100)Pu 238 μ Ci/L = [(Pu 238 dpm) $\times$ (DF) $\times$ (DDF) $\times$ (1000 mL/L)] / [(Pu-236 Tracer Recovery / 100) $\times$ (2220000 dpm/ μ Ci) $\times$ (D g/L) $\times$ (SS)]Relative Counting Error = Square Root of $\{[1 / (\text{Pu 236 cpm} \times \text{min})] + [1 / (\text{Pu 238 or 239/240 cpm} \times \text{min})]\} \times 1.96 \times 100$

02/03/99	Pu 239/240 μ Ci/mL	< 5.35E-07	DETECTION LEVELS in μ Ci/mL
02/02/99	Relative Counting Error	= 100.0%	Pu 239/240
03:05 PM	NOTE: Pu 238 Result is a LESS THAN Value.		5.35E-07
Sample Point	Pu 238 μ Ci/mL	< 5.35E-07	Pu 238
BY-105	Relative Counting Error	= 100.0%	5.35E-07
	Pu 236 Tracer Recovery	= 92.3%	

Analyst:	RRO	Date:	03-Feb-99
Signature of Chemist:	<i>John Ralyea</i>	JFR	3 Feb 99

SAMPLE WB1 REV 1.0

843128ML

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL28059-STD-PU
File ID: 7a7348.CNF

Counted on: 2/ 2/99 @22:58
Detector: AEA7
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1023.6	1023.6	363.188	363.188	10.000	2.695	5.000	0.862
2?	24.4	24.4	304.230	303.040	30.000	11.062	15.000	15.563
3	67.9	67.9	287.941	287.814	10.000	3.122	5.000	1.657
4?	20.5	20.5	270.763	269.954	10.000	3.618	5.000	0.816
5	1954.3	1954.3	229.456	229.455	10.000	2.957	5.000	1.279
6	12.8	12.8	144.218	144.192	138.000	1.000	69.000	0.100

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			Count Rate	%err	Activity	
			Exp.	Obs.	Diff.	c/m	@95	d/m uCi/ea	
1	Pu236	0.354	5.755	5.761	-0.0060.01	33.43	1.5	141.8 0.639E-04	
	Cm243		5.779	5.761	0.018			190.4 0.858E-04	
2		????		5.484		0.81	13.9		
3	Th228	0.018	5.400	5.414	-0.0140.01	1.68	7.5	9.8 0.442E-05	
4		????		5.332		0.90	11.6		
5	Pu239	0.564	5.147	5.146	0.0010.01	53.38	1.2	222.0 0.100E-03	
	Pu240		5.144	5.146	-0.002			222.0 0.100E-03	
6	Np237	0.011	4.769	4.754	0.0150.00	1.01	8.9	4.8 0.218E-05	
Totals:			0.946	<--valid peaks only-->			89.50		

DETECTOR CALIBRATION

Energy (MEV) = 4.090 + (0.0046)*Channel
Energy range (MeV): 4.090 TO 6.446
Efficiency = 0.2405 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	45393.0	100.000
Smoothed	45393.0	100.000
Composite fit	43779.5	96.446
Residuals	1613.5	3.554

Spectrum 7a7348.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 12828.5

6
6
6
6
6
6
6
6
6
6.
..
5..
...5..
.....5.
.....5
.....5
.....5
4.....5..
4
4
4
4
4
4
.3
3
2
2
2
1
..1
.....1
.....1.
.....1
.....1
.....1
1..

Raw Data Dump for AEA Spectrum: 7a7348.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	1.	4.
21	2.	5.	1.	0.	2.	1.	1.	1.	0.
31	4.	3.	4.	2.	2.	3.	3.	1.	1.
41	6.	0.	4.	0.	3.	2.	0.	2.	0.
51	1.	6.	2.	8.	2.	2.	3.	3.	1.
61	1.	1.	1.	4.	5.	3.	4.	2.	7.
71	0.	2.	3.	3.	2.	5.	6.	3.	3.
81	3.	4.	2.	3.	3.	3.	4.	5.	5.
91	4.	6.	2.	4.	1.	6.	4.	4.	9.
101	4.	8.	8.	3.	4.	1.	2.	4.	5.
111	4.	5.	7.	2.	5.	5.	6.	9.	9.
121	8.	5.	11.	6.	3.	4.	7.	8.	2.
131	3.	9.	8.	7.	4.	13.	9.	3.	4.
141	11.	12.	10.	14.	10.	12.	8.	7.	10.
151	13.	14.	16.	13.	9.	18.	15.	10.	13.
161	19.	15.	10.	13.	17.	16.	21.	18.	22.
171	19.	22.	21.	19.	27.	29.	16.	36.	35.
181	22.	34.	41.	45.	48.	43.	44.	38.	46.
191	57.	62.	66.	70.	71.	68.	77.	86.	96.
201	118.	113.	127.	124.	170.	189.	188.	217.	223.
211	284.	316.	325.	373.	471.	493.	547.	629.	725.
221	827.	858.	935.	974.	1176.	1355.	1743.	1958.	2219.
231	2188.	1501.	643.	202.	39.	13.	4.	7.	6.
241	4.	7.	9.	6.	9.	5.	15.	8.	4.
251	8.	7.	10.	9.	6.	6.	9.	18.	14.
261	20.	8.	12.	18.	13.	22.	15.	22.	33.
271	38.	23.	24.	23.	14.	19.	14.	23.	22.
281	38.	34.	39.	48.	42.	57.	72.	104.	67.
291	39.	26.	20.	18.	14.	19.	29.	23.	24.
301	20.	22.	21.	28.	24.	22.	25.	18.	8.
311	20.	16.	15.	15.	26.	23.	22.	28.	30.
321	25.	39.	21.	31.	28.	36.	46.	47.	48.
331	50.	60.	58.	82.	76.	74.	88.	95.	101.
341	110.	151.	156.	165.	196.	241.	248.	288.	297.
351	424.	507.	620.	652.	676.	613.	531.	565.	539.
361	843.	1044.	1209.	1294.	1068.	567.	183.	34.	6.
371	1.	0.	0.	0.	0.	1.	0.	2.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	1.	0.	0.	0.
421	1.	0.	0.	1.	0.	0.	0.	0.	1.
431	0.	0.	1.	1.	0.	0.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	0.
451	0.	0.	0.	0.	0.	0.	0.	0.	0.
461	0.	0.	0.	0.	0.	1.	2.	1.	1.
471	2.	1.	1.	4.	3.	5.	0.	4.	0.
481	4.	3.	1.	0.	1.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
G E N E R A L   A L P H A   E N E R G Y   A N A L Y S I S
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
WL28059-BLK-PU
File ID: 8a8327.CNF

Counted on: 2/ 2/99 @22:59
Detector: AEA8
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	763.6	763.6	361.764	361.764	10.000	3.095	5.000	0.694
2	54.3	54.3	287.426	287.294	12.000	3.861	6.000	1.384
3	18.5	18.5	268.919	268.338	10.000	5.746	5.000	2.231
4	9.3	9.3	245.482	245.068	10.000	5.316	5.000	1.168

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Exp.	Peak Centroid Obs.	Diff.	FWHM Rate	Count Rate	%err	d/m	Activity uCi/ea
1	Pu236	0.921	5.755	5.757	-.0020	0.01	32.92	1.6	148.3	0.668E-04
2	Th228	0.049	5.400	5.414	-.0140	0.02	1.75	7.8	10.9	0.490E-05
3		0.017		5.327	0.03		0.62	18.5	2.8	0.124E-05
4		0.012		5.220	0.02		0.44	17.3	1.9	0.871E-06
Totals:		1.000	<---valid peaks only-->				35.73			

DETECTOR CALIBRATION

Energy (MEV) = 4.092 + (0.0046)*Channel

Energy range (MeV): 4.092 TO 6.448

Efficiency = 0.2265 CPM/DPM

(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17152.0	100.000
Smoothed	17151.6	99.998
Composite fit	17152.7	100.004
Residuals	-0.7	-0.004

Analyzed by:

NEW

Spectrum 8a8327.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc Display Max.: 4378.0

```

4
4
4
4
4
4
3
23.
32.
....2.
12..
.1.
..1.
.....1
..... 1
..... 1
..... 1
.....1..
..... 1
.....1
.1...

```

Raw Data Dump for AEA Spectrum: 8a8327.CNF

1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	0.	0.	1.
21	0.	1.	1.	0.	1.	1.	1.	0.	1.
31	1.	0.	0.	0.	0.	0.	0.	0.	0.
41	0.	0.	1.	1.	0.	0.	1.	0.	0.
51	0.	1.	0.	1.	1.	1.	1.	0.	1.
61	2.	0.	0.	1.	1.	0.	0.	1.	0.
71	0.	0.	0.	1.	1.	0.	1.	1.	0.
81	1.	0.	0.	0.	0.	0.	0.	0.	1.
91	1.	0.	0.	0.	0.	2.	1.	0.	1.
101	1.	1.	0.	2.	0.	0.	1.	0.	3.
111	1.	0.	1.	0.	0.	0.	0.	1.	1.
121	2.	0.	1.	0.	1.	1.	2.	1.	2.
131	1.	0.	0.	1.	1.	1.	0.	0.	2.
141	0.	2.	1.	1.	1.	0.	1.	0.	1.
151	1.	1.	2.	2.	1.	0.	1.	2.	1.
161	2.	1.	2.	0.	2.	0.	0.	4.	2.
171	0.	0.	0.	0.	1.	2.	1.	4.	0.
181	1.	2.	2.	0.	3.	3.	0.	1.	4.
191	1.	1.	5.	0.	2.	2.	1.	3.	1.
201	1.	3.	6.	3.	6.	1.	1.	2.	1.
211	5.	2.	3.	4.	2.	1.	2.	9.	3.
221	7.	8.	4.	4.	6.	1.	7.	4.	6.
231	4.	5.	1.	6.	7.	8.	8.	9.	7.
241	10.	9.	8.	8.	16.	18.	8.	9.	8.
251	12.	12.	10.	7.	10.	7.	17.	11.	15.
261	13.	19.	14.	19.	20.	22.	30.	30.	40.
271	20.	25.	22.	18.	33.	36.	37.	35.	36.
281	34.	36.	29.	45.	60.	62.	78.	64.	70.
291	32.	24.	19.	16.	8.	19.	20.	20.	21.
301	25.	32.	34.	31.	27.	26.	32.	25.	24.
311	42.	26.	50.	38.	37.	48.	42.	45.	53.
321	51.	62.	77.	51.	77.	78.	109.	80.	88.
331	119.	102.	114.	150.	138.	139.	150.	155.	186.
341	212.	231.	233.	244.	258.	320.	305.	324.	394.
351	516.	571.	580.	554.	495.	503.	521.	499.	599.
361	775.	893.	891.	703.	384.	118.	36.	2.	0.
371	0.	0.	0.	2.	0.	0.	0.	1.	0.
381	1.	0.	0.	0.	0.	1.	0.	0.	0.
391	0.	0.	0.	0.	1.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	1.	0.	0.	0.
421	0.	1.	0.	0.	0.	0.	0.	0.	0.
431	0.	0.	0.	1.	1.	1.	0.	0.	0.
441	0.	0.	0.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	0.	0.	0.	0.	0.
461	0.	1.	0.	1.	0.	1.	3.	0.	1.
471	1.	1.	1.	2.	3.	3.	3.	4.	2.
481	2.	2.	0.	0.	0.	1.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	2.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2391-SAM-PU
File ID: 9a9238.CNF

Counted on: 2/ 2/99 @23: 0
Detector: AEA9
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	1101.5	1101.5	360.851	360.851	10.000	3.049	5.000	1.175
2	17.7	17.7	302.803	302.452	30.000	11.042	15.000	15.507
3	67.7	67.7	285.738	285.676	10.000	3.183	5.000	1.783
4	17.4	17.4	269.473	267.590	14.000	5.254	7.000	1.636

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid		Count Rate	%err	d/m	Activity
			Exp.	Obs. Diff.	c/m	@95		uCi/ea
1	Pu236	0.901	5.755	5.756 -0.0010	32.47	1.6	136.9	0.617E-04
2	Am241	0.016	5.479	5.487 -0.0080	0.59	15.1	2.6	0.116E-05
	Pu238		5.487	5.487 0.000			3.4	0.152E-05
3	Th228	0.045	5.400	5.410 -0.0100	1.63	7.3	9.5	0.429E-05
4		0.018		5.327 0.02	0.65	15.1	2.7	0.121E-05
Totals:			0.981	<--valid peaks only-->		35.34		

DETECTOR CALIBRATION

Energy (MEV) = 4.096 + (0.0046)*Channel
Energy range (MeV): 4.096 TO 6.451
Efficiency = 0.2420 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	17292.0	100.000
Smoothed	17292.0	100.000
Composite fit	16962.1	98.092
Residuals	329.9	1.908

Analyzed by: _____

NEW

Spectrum 9a9238.CNF

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.: 5721.6

```

4
4
4
4.
4.
....3
3.
2
2
1
.1
....1
..... 1
.....1....
.....1
.....1
1...

```

Raw Data Dump for AEA Spectrum:	9a9238.CNF	HNF-1663 REV. 0							
1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	1.	1.	0.	0.
21	1.	1.	1.	2.	0.	2.	0.	0.	0.
31	1.	2.	1.	1.	0.	2.	2.	1.	2.
41	1.	0.	2.	0.	1.	0.	1.	1.	2.
51	4.	0.	0.	2.	4.	1.	1.	0.	0.
61	0.	3.	0.	1.	0.	2.	0.	1.	1.
71	3.	0.	1.	0.	0.	1.	0.	0.	4.
81	4.	1.	1.	1.	1.	1.	2.	2.	0.
91	1.	2.	1.	2.	1.	3.	3.	1.	0.
101	1.	3.	2.	1.	2.	1.	0.	1.	2.
111	1.	2.	0.	0.	0.	2.	1.	0.	0.
121	4.	3.	3.	0.	2.	2.	3.	2.	3.
131	2.	1.	3.	4.	4.	2.	2.	3.	1.
141	1.	2.	2.	4.	2.	3.	2.	0.	0.
151	3.	2.	3.	2.	3.	3.	3.	2.	0.
161	4.	1.	1.	2.	0.	3.	1.	4.	1.
171	0.	2.	2.	3.	2.	1.	1.	1.	1.
181	2.	1.	1.	0.	2.	0.	2.	0.	5.
191	1.	3.	1.	1.	4.	1.	3.	1.	2.
201	4.	4.	9.	0.	6.	4.	2.	2.	3.
211	4.	3.	2.	1.	1.	0.	2.	1.	4.
221	3.	4.	6.	2.	6.	9.	4.	11.	3.
231	4.	3.	2.	5.	2.	2.	0.	5.	8.
241	6.	2.	6.	5.	6.	4.	8.	6.	1.
251	5.	5.	5.	0.	5.	11.	3.	14.	12.
261	8.	11.	11.	27.	19.	23.	25.	24.	21.
271	26.	19.	18.	14.	23.	16.	21.	17.	31.
281	51.	32.	66.	66.	84.	78.	80.	64.	35.
291	13.	4.	11.	9.	14.	6.	11.	18.	6.
301	13.	14.	20.	16.	10.	14.	6.	11.	10.
311	12.	8.	16.	11.	9.	18.	21.	14.	17.
321	22.	17.	18.	18.	23.	27.	32.	28.	35.
331	42.	52.	58.	63.	65.	86.	79.	95.	88.
341	130.	107.	167.	200.	208.	248.	269.	368.	411.
351	587.	635.	573.	539.	538.	605.	703.	802.	985.
361	1337.	1308.	929.	604.	231.	60.	13.	1.	0.
371	0.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	0.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	0.	1.	0.	0.	1.	0.	0.	1.	0.
431	0.	0.	0.	0.	0.	1.	0.	0.	0.
441	0.	0.	1.	0.	0.	0.	0.	0.	1.
451	0.	0.	0.	0.	1.	0.	0.	0.	1.
461	0.	0.	0.	0.	0.	0.	4.	1.	3.
471	2.	2.	3.	2.	4.	4.	9.	5.	3.
481	1.	0.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

222-S Analytical Laboratory
GENERAL ALPHA ENERGY ANALYSIS
Rev. 2.10

DATA REDUCTION REPORT

SAMPLE
S98T2391-DUP-PU
File ID: 10a1078.CNF

Counted on: 2/ 2/99 @23: 1
Detector: AEA10
Geometry number: 1
Count time: 28802. Sec

PEAK ANALYSIS

Peak ID	Peak height		Peak center		FWHM		Tau	
	Initial	Final	Initial	Final	Initial	Final	Initial	Final
1	536.7	536.7	360.292	360.292	10.000	3.337	5.000	0.653
2	30.7	30.7	285.668	285.443	10.000	3.223	5.000	0.582

PEAK RESULTS

Peak Error Limit: 30%

Peak ID	Isotope	AEA Frac	Peak Centroid			FWHM	Count Rate c/m	%err @95	d/m	Activity uCi/ea
			Exp.	Obs.	Diff.					
1	Pu236	0.965	5.755	5.745	0.0100	0.02	25.93	1.8	113.7	0.512E-04
2	Th228	0.059	5.400	5.401	-0.0010	0.01	1.57	8.5	9.5	0.429E-05
Totals:		1.024	<--valid peaks only-->				27.50			

DETECTOR CALIBRATION

Energy (MEV) = 4.088 + (0.0046) * Channel
Energy range (MeV): 4.088 TO 6.443
Efficiency = 0.2327 CPM/DPM
(Data reduction compression factor: 1.)

TOTAL COUNT DATA:

Item	Total	% Recovery
Raw spectrum	12894.0	100.000
Smoothed	12894.0	100.000
Composite fit	13201.0	102.381
Residuals	-307.0	-2.381

Analyzed by: _____

NEW

1 Legend: Raw = ... Modeled Peaks = 1,2,..., etc

Display Max.:

3468.1

[illegible]

Raw Data Dump for AEA Spectrum: 10a1078.CNF									
1	0.	0.	0.	0.	0.	0.	0.	0.	0.
11	0.	0.	0.	0.	0.	0.	1.	0.	0.
21	0.	1.	0.	0.	0.	0.	2.	0.	0.
31	2.	1.	0.	1.	0.	0.	1.	1.	1.
41	0.	0.	0.	0.	0.	2.	0.	2.	1.
51	0.	0.	1.	0.	1.	0.	0.	1.	1.
61	0.	1.	0.	2.	0.	2.	0.	1.	1.
71	1.	0.	1.	1.	0.	2.	1.	1.	0.
81	0.	0.	0.	0.	0.	0.	1.	0.	2.
91	0.	1.	0.	0.	1.	1.	0.	0.	1.
101	0.	0.	2.	0.	1.	2.	0.	0.	1.
111	1.	1.	0.	1.	1.	0.	0.	0.	1.
121	1.	1.	0.	2.	0.	1.	0.	3.	1.
131	0.	0.	0.	0.	2.	1.	0.	2.	1.
141	1.	2.	1.	0.	6.	2.	0.	0.	2.
151	2.	0.	2.	1.	0.	0.	3.	2.	2.
161	1.	2.	1.	1.	1.	0.	0.	0.	2.
171	1.	0.	1.	2.	4.	2.	1.	1.	5.
181	0.	3.	3.	1.	2.	3.	1.	0.	2.
191	1.	3.	0.	2.	0.	1.	2.	0.	3.
201	3.	3.	2.	1.	1.	2.	3.	3.	2.
211	4.	3.	5.	0.	5.	2.	2.	5.	3.
221	4.	5.	5.	4.	2.	5.	1.	3.	5.
231	5.	3.	5.	8.	1.	6.	4.	3.	4.
241	6.	8.	6.	7.	4.	7.	3.	5.	9.
251	9.	12.	9.	14.	8.	5.	5.	11.	16.
261	18.	15.	13.	14.	13.	22.	14.	33.	26.
271	20.	22.	19.	17.	24.	22.	30.	16.	19.
281	28.	45.	27.	42.	43.	47.	43.	40.	19.
291	18.	14.	19.	15.	17.	20.	14.	16.	26.
301	15.	29.	21.	24.	16.	25.	24.	25.	24.
311	20.	25.	34.	28.	38.	42.	32.	45.	57.
321	48.	49.	44.	62.	60.	68.	71.	85.	74.
331	92.	88.	113.	111.	127.	125.	140.	145.	162.
341	221.	203.	223.	192.	227.	260.	299.	320.	335.
351	419.	424.	427.	380.	362.	427.	391.	441.	537.
361	626.	588.	410.	190.	73.	9.	2.	0.	0.
371	1.	0.	0.	0.	0.	0.	0.	0.	0.
381	0.	0.	0.	0.	0.	0.	0.	1.	0.
391	0.	0.	0.	0.	0.	0.	0.	0.	0.
401	0.	0.	0.	0.	0.	0.	0.	0.	0.
411	0.	0.	0.	0.	0.	0.	0.	0.	0.
421	0.	0.	0.	1.	1.	0.	0.	0.	1.
431	1.	0.	0.	0.	0.	0.	0.	0.	0.
441	0.	1.	0.	1.	0.	1.	0.	0.	0.
451	0.	1.	1.	0.	0.	1.	1.	1.	0.
461	0.	2.	1.	0.	1.	0.	1.	1.	0.
471	0.	2.	0.	4.	5.	0.	0.	3.	1.
481	2.	1.	0.	0.	0.	0.	0.	0.	0.
491	0.	0.	0.	0.	0.	0.	0.	0.	0.
511	0.	0.							

HNF-1663 REV. 0

OPPORTUNISTIC ANALYSIS

APPENDIX A

HNF-1663 REV. 0

THIS PAGE INTENTIONALLY LEFT BLANK

Appendix A. Opportunistic Analyte Results.
8Y-105 (2)

CORE NUMBER: 246
SEGMENT #: 1

SEGMENT PORTION: L Lower Half of Segment

Sample #	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err %
S981002598	Ag/g	100.0	<1.00e-02	<2.01e+02	<2.01e+02	N/A	N/A	92.80	201.0	N/A	N/A
S981002598	Ag/g	103.4	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	102.0	2.01e+03	N/A	N/A
S981002598	Ag/g	102.4	<5.00e-03	<1.01e+02	<1.01e+02	N/A	N/A	98.40	100.0	N/A	N/A
S981002598	Ag/g	101.6	<5.00e-03	<1.01e+02	<1.01e+02	N/A	N/A	100.0	100.0	N/A	N/A
S981002598	Ag/g	102.2	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	98.40	2.01e+03	N/A	N/A
S981002598	Ag/g	99.80	<1.00e-02	<4.03e+02	<4.03e+02	N/A	N/A	97.90	403.0	N/A	N/A
S981002598	Ag/g	96.40	<1.00e-02	<2.01e+02	<2.01e+02	N/A	N/A	92.50	201.0	N/A	N/A
S981002598	Ag/g	101.4	<5.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	97.00	1.00e+03	N/A	N/A
S981002598	Ag/g	102.6	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	99.00	1.00e+03	N/A	N/A
S981002598	Ag/g	100.8	<5.00e-02	<1.01e+03	<1.01e+03	N/A	N/A	99.00	1.00e+03	N/A	N/A
S981002598	Ag/g	99.40	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	94.80	2.01e+03	N/A	N/A
S981002598	Ag/g	100.6	<2.00e-01	<4.03e+02	<4.03e+02	N/A	N/A	99.70	4.03e+02	N/A	N/A
S981002598	Ag/g	99.60	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	99.70	2.01e+03	N/A	N/A
S981002598	Ag/g	101.6	<6.00e-02	<1.01e+03	<1.01e+03	N/A	N/A	99.10	1.01e+03	N/A	N/A
S981002598	Ag/g	96.00	<1.00e-01	<2.01e+03	<2.01e+03	N/A	N/A	95.80	2.01e+03	N/A	N/A
S981002598	Ag/g	99.40	<1.00e-02	<2.01e+02	<2.01e+02	N/A	N/A	94.50	2.01e+03	N/A	N/A
S981002598	Ag/g	101.6	<1.00e-02	<2.01e+02	<2.01e+02	N/A	N/A	97.20	201.0	N/A	N/A
S981002598	Ag/g	96.00	<2.00e-01	<4.03e+02	<4.03e+02	N/A	N/A	93.50	4.03e+02	N/A	N/A
S981002598	Ag/g	100.2	<5.00e-02	<1.01e+03	<1.01e+03	N/A	N/A	97.90	1.00e+03	N/A	N/A
S981002598	Ag/g	105.3	<4.85e-03	<5.21e+03	<4.83e+03	N/A	N/A	3.79e+06	5.554	N/A	N/A
S981002604	Ag/g	115.3	7.00e-03	<5.555	<5.380	N/A	N/A	3.64e+06	7.40e-02	N/A	N/A
S981002604	Ag/g	97.00	<2.00e-05	1.26e-01	7.90e-02	N/A	N/A	7.40e-02	7.40e-02	N/A	N/A
S981002604	Ag/g	N/A	2.53e-05	<7.41e-02	<7.18e-2	N/A	N/A	N/A	7.40e-02	N/A	N/A
S981002604	Ag/g	N/A	4.18e-05	<7.41e-02	<7.18e-2	N/A	N/A	N/A	7.40e-02	N/A	N/A
S981002604	Ag/g	N/A	<3.53e-05	<1.19e-01	<1.15e-1	N/A	N/A	N/A	1.19e-01	N/A	N/A
S981002604	Ag/g	N/A	<3.51e-05	<1.19e-01	<1.15e-1	N/A	N/A	N/A	1.19e-01	N/A	N/A
S981002604	Ag/g	108.0	<5.08e-05	<2.13e-01	<1.08e-1	N/A	N/A	101.5	1.71e-01	N/A	N/A
S981002604	Ag/g	103.9	<3.59e-05	<2.13e-01	<1.17e-1	N/A	N/A	103.3	1.21e-01	N/A	N/A
S981002604	Ag/g	N/A	<4.47e-05	<1.50e-01	<1.46e-1	N/A	N/A	N/A	1.50e-01	N/A	N/A
S981002604	Ag/g	N/A	<6.47e-05	<2.18e-01	<2.11e-1	N/A	N/A	N/A	2.18e-01	N/A	N/A
S981002604	Ag/g	14.60	<1.00e-02	<9.40	18.30	N/A	N/A	15.10	9.200	N/A	N/A
S981002604	Ag/g	98.60	<1.00e-01	<9.210	<8.93e-1	N/A	N/A	102.8	92.00	N/A	N/A
S981002604	Ag/g	96.00	<5.00e-03	<4.610	<4.42e-0	N/A	N/A	97.60	4.610	N/A	N/A
S981002604	Ag/g	96.00	<1.00e-01	<9.210	<8.93e-0	N/A	N/A	100.8	4.610	N/A	N/A
S981002604	Ag/g	94.80	<2.00e-02	<18.40	<1.79e-1	N/A	N/A	98.00	18.40	N/A	N/A
S981002604	Ag/g	91.00	<1.00e-02	<9.210	<8.93e-0	N/A	N/A	97.40	9.200	N/A	N/A
S981002604	Ag/g	97.80	<5.00e-02	<46.10	<4.42e-1	N/A	N/A	97.00	46.10	N/A	N/A
S981002604	Ag/g	97.60	<1.00e-01	<9.210	<8.93e-1	N/A	N/A	99.80	92.00	N/A	N/A
S981002604	Ag/g	97.20	<5.00e-02	<46.10	<4.42e-1	N/A	N/A	100.8	46.10	N/A	N/A
S981002604	Ag/g	95.60	<1.00e-01	<9.210	<8.93e-1	N/A	N/A	94.20	92.00	N/A	N/A

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002604	A	Phosphorus -ICP-Acid Digest	µg/g	95.40	< 2.00e-01	< 1.85e-02	< 1.79e-02	N/A	N/A	92.40	182.0	N/A
S981002604	A	Lead -ICP-Acid Digest	µg/g	95.40	< 1.00e-01	< 9.2-10	< 8.93e-1	N/A	N/A	101.8	92.00	N/A
S981002604	A	Sulfur -ICP-Acid Digest	µg/g	92.00	< 1.00e-01	< 9.2-10	< 8.93e-1	N/A	N/A	103.2	92.00	N/A
S981002604	A	Antimony -ICP-Acid Digest	µg/g	95.40	< 4.00e-02	< 5.5-30	< 5.32e-1	N/A	N/A	100.2	55.50	N/A
S981002604	A	Selenium -ICP-Acid Digest	µg/g	92.00	< 1.00e-01	< 9.2-10	< 8.93e-1	N/A	N/A	97.20	92.00	N/A
S981002604	A	Samarium -ICP-Acid Digest	µg/g	92.40	< 1.00e-02	< 9.2-10	< 8.93e-1	N/A	N/A	95.60	92.00	N/A
S981002604	A	Yttrium -ICP-Acid Digest	µg/g	95.40	< 1.00e-02	< 9.2-10	< 8.93e-1	N/A	N/A	96.60	9.200	N/A
S981002604	A	Thallium -ICP-Acid Digest	µg/g	95.40	< 2.00e-01	< 1.85e-02	< 1.79e-02	N/A	N/A	98.80	182.0	N/A
S981002604	A	Vanadium -ICP-Acid Digest	µg/g	95.40	< 5.00e-02	< 6.6-10	< 6.44e-1	N/A	N/A	99.80	46.10	N/A
S981002610	W	Recreate by IC-DIONEY 4000/4500	µg/g	79.58	< 2.00e-02	< 22.62	< 28.30	N/A	N/A	79.50	22.63	N/A
S981002610	W	Glycolate IC-DIONEY 4000/4500	µg/g	94.53	< 5.60e-02	< 63.35	< 65.31e1	N/A	N/A	90.52	63.36	N/A

25-mar-1999 09:46:13
A-0002-1CODE NUMBER: 246
SEGMENT #: 2

SEGMENT POSITION: L Lower Half of Segment

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

Sample	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002599	F	Silver - ICP-Fusion	ug/g	100.0	<1.00e-02	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Arsenic - ICP-Fusion	ug/g	103.4	<1.00e-01	<1.99e-03	<1.99e-03	n/a	n/a	1.99e-03	n/a
S98T002599	F	Beryllium - ICP-Fusion	ug/g	101.6	<3.00e-03	< 99.50	< 99.50	n/a	n/a	99.60	n/a
S98T002599	F	Cadmium - ICP-Fusion	ug/g	101.6	<3.00e-03	< 99.50	< 99.50	n/a	n/a	99.60	n/a
S98T002599	F	Cobalt - ICP-Fusion	ug/g	102.4	<1.00e-01	<1.99e-03	<1.99e-03	n/a	n/a	1.99e-03	n/a
S98T002599	F	Copper - ICP-Fusion	ug/g	99.60	<2.00e-02	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Lead - ICP-Fusion	ug/g	101.6	<1.00e-02	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Lithium - ICP-Fusion	ug/g	102.4	<3.00e-03	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Magnesium - ICP-Fusion	ug/g	101.6	<3.00e-03	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Molybdenum - ICP-Fusion	ug/g	100.8	<3.00e-03	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Neodymium - ICP-Fusion	ug/g	99.60	<1.00e-01	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Phosphorus - ICP-Fusion	ug/g	100.4	<2.00e-01	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Lead - ICP-Fusion	ug/g	99.60	<1.00e-01	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Selenium - ICP-Fusion	ug/g	97.80	<1.00e-02	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Antimony - ICP-Fusion	ug/g	101.6	<3.00e-03	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Strontium - ICP-Fusion	ug/g	102.4	<1.00e-02	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Titanium - ICP-Fusion	ug/g	101.6	<1.00e-02	<1.99e-02	<1.99e-02	n/a	n/a	199.0	n/a
S98T002599	F	Thallium - ICP-Fusion	ug/g	96.00	<2.00e-01	<3.98e-02	<3.98e-02	n/a	n/a	398.0	n/a
S98T002599	F	Vannadium - ICP-Fusion	ug/g	100.2	<5.00e-02	<9.97e-02	<9.97e-02	n/a	n/a	99.60	n/a
S98T002599	F	Cobalt-60 by GE	uCi/g	103.8	<4.68e-02	<4.68e-03	<4.68e-03	n/a	n/a	4.68e-03	n/a
S98T002599	A	Sn117 ICP/MS Acid Dig LA-505159	ug/g	115.3	7.00e-03	< 5.353	< 5.353	n/a	n/a	5.353	n/a
S98T002605	A	Ce133 ICP/MS Acid Dig LA-505159	ug/g	97.00	<2.20e-05	2.79e-01	3.09e-01	2.92e-01	11.4	n/a	7.10e-02
S98T002605	A	Ce135 ICP/MS Acid Dig LA-505159	ug/g	n/a	2.53e-05	8.92e-02	9.94e-02	9.43e-02	10.8	n/a	7.10e-02
S98T002605	A	Ce137 ICP/MS Acid Dig LA-505159	ug/g	n/a	4.18e-05	9.99e-02	1.22e-01	1.11e-01	19.9	n/a	7.10e-02
S98T002605	A	Protactinium-231 by ICP/MS	ug/g	n/a	<3.51e-05	<1.18e-01	<1.18e-01	n/a	n/a	1.18e-01	n/a
S98T002605	A	Thorium-229 by ICP/MS	ug/g	n/a	<3.51e-05	<1.18e-01	<1.18e-01	n/a	n/a	1.18e-01	n/a
S98T002605	A	Thorium-230 by ICP/MS	ug/g	n/a	<3.51e-05	<1.18e-01	<1.18e-01	n/a	n/a	1.18e-01	n/a
S98T002605	A	Pu/Am-241 by ICP/MS	ug/g	108.0	<3.59e-05	<1.68e-01	<1.70e-01	n/a	n/a	1.68e-01	n/a
S98T002605	A	Am 243/Am 244 by ICP/MS	ug/g	103.9	<3.59e-05	<1.68e-01	<1.70e-01	n/a	n/a	1.68e-01	n/a
S98T002605	A	Pu 243/Am 244 by ICP/MS	ug/g	n/a	<4.47e-05	<1.45e-01	<1.50e-01	n/a	n/a	1.45e-01	n/a
S98T002605	A	Silver - ICP-Acid Digest	ug/g	14.60	<1.00e-02	17.40	19.30	18.35	10.4	n/a	8.900
S98T002605	A	Arsenic - ICP-Acid Digest	ug/g	98.60	<1.00e-02	< 88.80	< 88.80	< 88.80	n/a	89.00	n/a
S98T002605	A	Beryllium - ICP-Acid Digest	ug/g	97.20	<5.00e-03	< 4.440	< 4.440	< 4.440	n/a	4.440	n/a
S98T002605	A	Cadmium - ICP-Acid Digest	ug/g	96.00	<5.00e-03	< 4.440	< 4.440	< 4.440	n/a	4.440	n/a
S98T002605	A	Cerium - ICP-Acid Digest	ug/g	98.60	<1.00e-01	< 88.80	< 88.80	< 88.80	n/a	89.00	n/a
S98T002605	A	Cobalt - ICP-Acid Digest	ug/g	94.80	<2.00e-02	< 17.80	< 17.80	< 17.80	n/a	17.80	n/a
S98T002605	A	Copper - ICP-Acid Digest	ug/g	91.00	<1.00e-02	< 8.880	< 8.880	< 8.880	n/a	8.900	n/a
S98T002605	A	Lanthanum - ICP-Acid Digest	ug/g	97.80	<5.00e-02	< 44.40	< 44.40	< 44.40	n/a	44.40	n/a
S98T002605	A	Magnesium - ICP-Acid Digest	ug/g	97.80	<1.00e-01	< 88.80	< 88.80	< 88.80	n/a	89.00	n/a
S98T002605	A	Molybdenum - ICP-Acid Digest	ug/g	97.60	<5.00e-02	< 44.40	< 44.40	< 44.40	n/a	44.40	n/a
S98T002605	A	Neodymium - ICP-Acid Digest	ug/g	95.60	<1.00e-01	< 88.80	< 88.80	< 88.80	n/a	89.00	n/a

5052

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec. %	Det Limit	Count Err%
S981002605	A	Phosphorus -ICP-Acid Digest	ug/g	95.40	<2.00e-01	1.85e-02	201.0	193.0	8.29	N/A	178.0	N/A
S981002605	A	Lead -ICP-Acid Digest	ug/g	95.40	<1.00e-01	8.8e-02	-9.14e1	N/A	N/A	N/A	89.00	N/A
S981002605	A	Sulfur -ICP-Acid Digest	ug/g	92.00	<1.00e-01	2.83e-02	303.0	295.0	5.42	N/A	89.00	N/A
S981002605	A	Antimony -ICP-Acid Digest	ug/g	95.40	<6.00e-02	8.5e-02	-5.50e1	N/A	N/A	N/A	53.50	N/A
S981002605	A	Selenium -ICP-Acid Digest	ug/g	92.00	<1.00e-01	88.80	-9.14e1	N/A	N/A	N/A	89.00	N/A
S981002605	A	Samarium -ICP-Acid Digest	ug/g	96.30	<1.00e-01	88.80	-9.14e1	N/A	N/A	N/A	89.00	N/A
S981002605	A	Strontium -ICP-Acid Digest	ug/g	97.60	<1.00e-02	8.880	-9.14e0	N/A	N/A	N/A	8.900	N/A
S981002605	A	Titanium -ICP-Acid Digest	ug/g	98.00	<1.00e-02	8.880	-9.14e0	N/A	N/A	N/A	8.900	N/A
S981002605	A	Thallium -ICP-Acid Digest	ug/g	92.00	<2.00e-01	41.78e-02	-1.83e-2	N/A	N/A	N/A	178.0	N/A
S981002605	A	Vanadium -ICP-Acid Digest	ug/g	98.00	<5.00e-02	66.40	-46.58e1	N/A	N/A	N/A	44.40	N/A
S981002611	W	Acetate by IC-DIONE	4000/4500	79.58	<2.00e-02	60.90	57.40	59.17	5.92	N/A	23.00	N/A
S981002611	W	Glycolate IC-Dione	4000/4500	94.83	<5.00e-02	64.39	-66.65e1	N/A	N/A	N/A	64.38	N/A

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 246
SEGMENT #: 3

SEGMENT PORTION: 1 Lower Half of Segment

Sample #	Alt #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count	Err%
S981002600	F	Stilver - ICP-Fusion	ug/g	103.0	< 1.00e-02	2.03e-02	< 2.04e-2	N/A	N/A	N/A	203.0	N/A
S981002600	F	Arsenic - ICP-Fusion	ug/g	103.4	< 1.00e-03	2.03e-03	< 2.04e-3	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Beryllium - ICP-Fusion	ug/g	102.4	< 2.00e-01	4.05e-02	< 4.05e-2	N/A	N/A	N/A	102.0	N/A
S981002600	F	Cadmium - ICP-Fusion	ug/g	101.5	< 1.00e-01	2.03e-02	< 2.04e-2	N/A	N/A	N/A	102.0	N/A
S981002600	F	Cerium - ICP-Fusion	ug/g	102.2	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Cobalt - ICP-Fusion	ug/g	99.60	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	203.0	N/A
S981002600	F	Copper - ICP-Fusion	ug/g	98.40	< 1.00e-02	2.03e-02	< 2.04e-2	N/A	N/A	N/A	203.0	N/A
S981002600	F	Lead - ICP-Fusion	ug/g	102.4	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	1.02e+03	N/A
S981002600	F	Mercury - ICP-Fusion	ug/g	101.8	< 1.00e-02	2.03e-02	< 2.04e-2	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Magnesium - ICP-Fusion	ug/g	101.8	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	1.02e+03	N/A
S981002600	F	Molybdenum - ICP-Fusion	ug/g	99.60	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Nickel - ICP-Fusion	ug/g	100.2	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	4.07e+03	N/A
S981002600	F	Phosphorus - ICP-Fusion	ug/g	99.60	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Silver - ICP-Fusion	ug/g	97.80	< 1.00e-01	6.88e-03	< 7.17e-03	7.02e+03	4.13	N/A	2.03e+03	N/A
S981002600	F	Antimony - ICP-Fusion	ug/g	101.5	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	1.22e+03	N/A
S981002600	F	Selenium - ICP-Fusion	ug/g	96.00	< 1.00e-01	2.03e+03	< 2.04e-3	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Strontium - ICP-Fusion	ug/g	101.5	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	2.03e+03	N/A
S981002600	F	Thallium - ICP-Fusion	ug/g	102.4	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	203.0	N/A
S981002600	F	Vanadium - ICP-Fusion	ug/g	96.00	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	4.07e+03	N/A
S981002600	F	Zinc - ICP-Fusion	ug/g	100.2	< 1.00e-02	4.05e-02	< 4.05e-2	N/A	N/A	N/A	5.00e-03	N/A
S981002600	F	Chromium - ICP-Fusion	ug/g	103.8	< 4.05e-03	5.14e-03	< 5.14e-3	N/A	N/A	N/A	5.00e-03	N/A
S981002600	F	Cadmium - ICP-Fusion	ug/g	115.3	< 7.00e-03	6.59e-03	< 6.59e-3	N/A	N/A	N/A	5.652	N/A
S981002600	A	Sn117 ICP/MS Acid/LA-505159	ug/g	97.00	< 2.00e-05	5.39e-01	< 5.39e-1	5.87e-01	16.4	N/A	7.50e-02	N/A
S981002600	A	Cs133 ICP/MS Acid/LA-505159	ug/g	97.00	< 2.00e-05	3.43e-01	< 3.43e-1	3.64e-01	11.8	N/A	7.50e-02	N/A
S981002600	A	Cs137 ICP/MS Acid/LA-505159	ug/g	N/A	< 2.00e-05	9.95e-01	< 9.95e-1	1.072	14.5	N/A	7.50e-02	N/A
S981002600	A	Protactinium-231 by ICP/MS	ug/g	N/A	< 4.18e-05	1.19e-01	< 1.19e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Thorium-230 by ICP/MS	ug/g	N/A	< 3.51e-05	< 1.20e-01	< 1.20e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Thorium-232 by ICP/MS	ug/g	N/A	< 3.29e-05	< 1.13e-01	< 1.13e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Pu239/241 by ICP/MS	ug/g	103.0	< 5.09e-05	6.068	< 6.068	6.594	15.9	N/A	1.20e-01	N/A
S981002600	A	Pu239/241 by ICP/MS	ug/g	103.9	< 5.09e-05	< 2.28e-01	< 2.28e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Pu239/241 by ICP/MS	ug/g	N/A	< 4.47e-05	< 2.28e-01	< 2.28e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Pu239/241 by ICP/MS	ug/g	N/A	< 4.47e-05	< 2.28e-01	< 2.28e-1	N/A	N/A	N/A	1.20e-01	N/A
S981002600	A	Protactinium-231 by ICP/MS	ug/g	14.60	< 1.00e-02	< 93.70	< 93.70	21.25	4.24	N/A	9.350	N/A
S981002600	A	Arsenic - ICP-Acid Digest	ug/g	98.60	< 1.00e-03	< 8.870	< 8.870	9.400	11.3	N/A	4.650	N/A
S981002600	A	Beryllium - ICP-Acid Digest	ug/g	96.00	< 5.00e-03	< 93.70	< 93.70	N/A	N/A	N/A	93.50	N/A
S981002600	A	Cadmium - ICP-Acid Digest	ug/g	96.00	< 1.00e-02	< 18.70	< 18.70	N/A	N/A	N/A	18.80	N/A
S981002600	A	Cerium - ICP-Acid Digest	ug/g	98.60	< 1.00e-02	< 9.370	< 9.370	N/A	N/A	N/A	9.350	N/A
S981002600	A	Cobalt - ICP-Acid Digest	ug/g	97.80	< 1.00e-02	< 46.90	< 46.90	N/A	N/A	N/A	46.80	N/A
S981002600	A	Copper - ICP-Acid Digest	ug/g	97.80	< 1.00e-02	< 93.70	< 93.70	N/A	N/A	N/A	93.50	N/A
S981002600	A	Lead - ICP-Acid Digest	ug/g	97.60	< 1.00e-02	< 46.90	< 46.90	N/A	N/A	N/A	46.80	N/A
S981002600	A	Magnesium - ICP-Acid Digest	ug/g	97.60	< 1.00e-02	< 93.70	< 93.70	N/A	N/A	N/A	93.50	N/A
S981002600	A	Molybdenum - ICP-Acid Digest	ug/g	95.60	< 1.00e-01	< 93.70	< 93.70	N/A	N/A	N/A	93.50	N/A
S981002600	A	Nickel - ICP-Acid Digest	ug/g	95.60	< 1.00e-01	< 93.70	< 93.70	N/A	N/A	N/A	93.50	N/A

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002606	A	Phosphorus -IC-Acid Digest	ug/g	95.40	<2.00e-01	3.45e-03	3.21e-03	3.44e-03	11.9	N/A	185.0
S981002606	A	Lead -IC-Acid Digest	ug/g	95.20	<1.00e-01	<3.93e-03	<9.25e-03	N/A	N/A	N/A	N/A
S981002606	A	Sulfur -IC-Acid Digest	ug/g	92.00	<1.00e-01	5.03e-03	6.01e-03	5.52e-03	17.8	N/A	93.50
S981002606	A	Antimony -IC-Acid Digest	ug/g	95.40	<6.00e-02	<8.5e-03	<5.54e-03	N/A	N/A	N/A	56.00
S981002606	A	Selenium -IC-Acid Digest	ug/g	92.00	<1.00e-01	93.70	<9.23e-03	N/A	N/A	N/A	93.50
S981002606	A	Samarium -IC-Acid Digest	ug/g	96.00	<1.00e-01	93.70	<9.23e-03	N/A	N/A	N/A	93.50
S981002606	A	Strontium -IC-Acid Digest	ug/g	97.00	<1.00e-02	1.03e-02	<9.11e-03	109.0	11.0	N/A	9.350
S981002606	A	Titanium -IC-Acid Digest	ug/g	98.00	<1.00e-02	<9.37e-03	<9.22e-03	N/A	N/A	N/A	9.350
S981002606	A	Thallium -IC-Acid Digest	ug/g	92.00	<2.00e-01	<1.87e-02	<1.85e-02	N/A	N/A	N/A	188.0
S981002606	A	Vanadium -IC-Acid Digest	ug/g	96.00	<5.00e-02	<4.6e-02	<4.62e-02	N/A	N/A	N/A	46.80
S981002612	W	Acetate by IC-DIONEK 4000/4500	ug/g	96.55	<2.00e-02	2.36e-02	309.0	269.6	29.3	N/A	22.90
S981002612	W	Glycolate IC-DIONEK 4000/4500	ug/g	94.83	<5.60e-02	<6.11e-02	<6.43e-02	N/A	N/A	N/A	64.08

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err %
S98T002601	F	Si/Alver - ICP-Fusion	ug/g	100.2	<1.00e-02	<1.99e-02	<1.99e-2	N/A	N/A	N/A	198.0	N/A
S98T002601	F	Arsenic - ICP-Fusion	ug/g	104.2	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Beryllium - ICP-Fusion	ug/g	103.2	<5.00e-03	99.30	<9.9e-1	N/A	N/A	N/A	99.20	N/A
S98T002601	F	Cadmium - ICP-Fusion	ug/g	102.2	<5.00e-03	99.30	<9.9e-1	N/A	N/A	N/A	99.20	N/A
S98T002601	F	Cerium - ICP-Fusion	ug/g	102.8	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Cobalt - ICP-Fusion	ug/g	102.2	<2.00e-02	<3.97e-02	<3.98e-2	N/A	N/A	N/A	397.0	N/A
S98T002601	F	Copper - ICP-Fusion	ug/g	96.80	<1.00e-02	<1.99e-02	<1.99e-2	N/A	N/A	N/A	198.0	N/A
S98T002601	F	Lanthanum - ICP-Fusion	ug/g	102.2	<5.00e-02	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Magnesium - ICP-Fusion	ug/g	100.4	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Molybdenum - ICP-Fusion	ug/g	101.4	<5.00e-02	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Niodymium - ICP-Fusion	ug/g	100.4	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	3.97e-03	N/A
S98T002601	F	Phosphorus - ICP-Fusion	ug/g	100.4	<2.00e-01	<1.97e-03	<3.98e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Lead - ICP-Fusion	ug/g	100.4	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Sulfur - ICP-Fusion	ug/g	97.60	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Antimony - ICP-Fusion	ug/g	101.4	<5.00e-02	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Selenium - ICP-Fusion	ug/g	98.00	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Strontium - ICP-Fusion	ug/g	100.2	<1.00e-01	<1.99e-03	<1.99e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Titanium - ICP-Fusion	ug/g	102.6	<1.00e-02	<1.99e-02	<1.99e-2	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Thallium - ICP-Fusion	ug/g	103.0	<1.00e-02	<1.99e-02	<1.99e-2	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Vanadium - ICP-Fusion	ug/g	98.40	<2.00e-01	<3.97e-03	<3.98e-3	N/A	N/A	N/A	1.98e-03	N/A
S98T002601	F	Vanadium - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.99e-02	<1.99e-2	N/A	N/A	N/A	1.98e-02	N/A
S98T002601	F	Cobalt-60 by BGA	uCi/g	108.7	<2.99e-05	<1.33e-02	<1.30e-2	N/A	N/A	N/A	1.98e-02	N/A
S98T002607	A	Sm171 ICP/MS Actidig LA-505159	ug/g	115.3	3.00e-03	5.854	46.02e0	N/A	N/A	N/A	5.853	N/A
S98T002607	A	Cs133 ICP/MS Actidig LA-505159	ug/g	99.00	2.00e-05	3.00e-01	2.85e-01	2.85e-01	10.9	N/A	7.80e-02	N/A
S98T002607	A	Cs135 ICP/MS Actidig LA-505159	ug/g	N/A	5.00e-05	1.7e-01	1.00e-01	1.09e-01	15.7	N/A	7.80e-02	N/A
S98T002607	A	Cs137 ICP/MS Actidig LA-505159	ug/g	N/A	8.50e-05	3.2e-01	2.95e-01	3.08e-01	10.0	N/A	7.80e-02	N/A
S98T002607	A	Protactinium-231 by ICP/MS	ug/g	N/A	<5.57e-05	<2.17e-01	<2.24e-1	N/A	N/A	N/A	2.17e-01	N/A
S98T002607	A	Thorium-229 by ICP/MS	ug/g	N/A	<4.46e-05	<1.74e-01	<1.79e-1	N/A	N/A	N/A	1.74e-01	N/A
S98T002607	A	Thorium-230 by ICP/MS	ug/g	N/A	<3.59e-05	<1.39e-01	<1.43e-1	N/A	N/A	N/A	1.39e-01	N/A
S98T002607	A	Thorium-232 by ICP/MS	ug/g	102.2	<1.92e-05	1.456	1.360	1.408	7.09	N/A	7.50e-02	N/A
S98T002607	A	Pu/Am-241 by ICP/MS	ug/g	110.1	<3.71e-05	<1.45e-01	<1.49e-1	N/A	N/A	N/A	1.45e-01	N/A
S98T002607	A	Pu/Am-241 by ICP/MS	ug/g	N/A	<3.83e-05	<1.50e-01	<1.54e-1	N/A	N/A	N/A	1.50e-01	N/A
S98T002607	A	Pu/Am-241 by ICP/MS	ug/g	N/A	<3.41e-05	<1.33e-01	<1.37e-1	N/A	N/A	N/A	1.33e-01	N/A
S98T002607	A	Pu/Am-241 by ICP/MS	ug/g	82.00	<1.00e-02	19.30	19.30	19.20	1.04	N/A	7.60	N/A
S98T002607	A	Si/Ver - ICP-Acid Digest	ug/g	90.60	<1.00e-01	<7.77e-01	<7.99e1	N/A	N/A	N/A	77.60	N/A
S98T002607	A	Arsenic - ICP-Acid Digest	ug/g	101.8	<5.00e-03	<3.880	<4.00e0	N/A	N/A	N/A	3.880	N/A
S98T002607	A	Beryllium - ICP-Acid Digest	ug/g	88.40	<5.00e-03	<3.880	<4.00e0	N/A	N/A	N/A	3.880	N/A
S98T002607	A	Cadmium - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<7.77e-01	<7.99e1	N/A	N/A	N/A	77.60	N/A
S98T002607	A	Cerium - ICP-Acid Digest	ug/g	89.20	<2.00e-02	<15.50	<1.60e1	N/A	N/A	N/A	15.50	N/A
S98T002607	A	Cobalt - ICP-Acid Digest	ug/g	90.20	<1.00e-02	<7.77e-01	<7.99e0	N/A	N/A	N/A	7.760	N/A
S98T002607	A	Lanthanum - ICP-Acid Digest	ug/g	98.00	<5.00e-02	<36.80	<4.00e1	N/A	N/A	N/A	36.80	N/A
S98T002607	A	Copper - ICP-Acid Digest	ug/g	93.80	<1.00e-01	<7.77e-01	<7.99e1	N/A	N/A	N/A	77.60	N/A
S98T002607	A	Magnesium - ICP-Acid Digest	ug/g	92.40	<5.00e-02	<36.80	<4.00e1	N/A	N/A	N/A	36.80	N/A
S98T002607	A	Molybdenum - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<7.77e-01	<7.99e1	N/A	N/A	N/A	77.60	N/A
S98T002607	A	Niodymium - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<7.77e-01	<7.99e1	N/A	N/A	N/A	77.60	N/A

5056

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002607	A	Phosphorus -ICP-Acid Digest	ug/g	90.60	<2.00e-01	1.10e-03	1.11e-03	1.11e-03	1.80	D/A	155.0
S981002607	A	Lead -ICP-Acid Digest	ug/g	82.20	<1.00e-01	<7.70	<7.69e-01	<7.69e-01	D/A	77.60	D/A
S981002607	A	Sulfur -ICP-Acid Digest	ug/g	88.40	<1.00e-01	5.90e-02	4.79e-02	5.42e-02	17.5	D/A	77.60
S981002607	A	Antimony -ICP-Acid Digest	ug/g	88.60	<6.00e-02	<4.70	<4.70e-01	<4.70e-01	D/A	46.80	D/A
S981002607	A	Selenium -ICP-Acid Digest	ug/g	88.60	<1.00e-01	<7.70	<7.69e-01	<7.69e-01	D/A	77.60	D/A
S981002607	A	Samarium -ICP-Acid Digest	ug/g	97.40	<1.00e-01	<7.70	<7.69e-01	<7.69e-01	D/A	77.60	D/A
S981002607	A	Strontium -ICP-Acid Digest	ug/g	97.40	<1.00e-02	13.50	<7.69e-02	12.80	9.30	D/A	7.760
S981002607	A	Titanium -ICP-Acid Digest	ug/g	96.90	<1.00e-02	<7.70	<7.69e-02	<7.69e-02	D/A	7.760	D/A
S981002607	A	Thallium -ICP-Acid Digest	ug/g	86.80	<2.00e-01	<4.55e-02	<4.60e-02	D/A	D/A	155.0	D/A
S981002607	A	Vanadium -ICP-Acid Digest	ug/g	96.80	<5.00e-02	<38.80	<4.00e-01	D/A	D/A	38.80	D/A
S981002613	W	Acetate by IC-DIONE	4000/4500	100.0	<2.00e-02	2.32e-02	2.13e-02	2.26e-02	11.5	D/A	84.92
S981002613	W	Glycolate IC-DIONE	4000/4500	100.0	<5.00e-02	<2.36e-02	<2.33e-02	D/A	D/A	237.7	D/A

25-mar-1999 09:56:10
A-0002-1

CORE NUMBER: 246
SEGMENT #: 8

Appendix A. Opportunistic Analyte Results. 81-105 (2)

SEGMENT PORTION: L Lower Half of Segment											
Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002712	F	Silver - ICP-Fusion	ug/g	100.4	< 1.00e-02	< 2.08e+02	< 2.09e2	n/a	n/a	n/a	208.0
S981002712	F	Silver - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Arsenic - ICP-Fusion	ug/g	103.4	< 5.00e-03	< 1.04e+02	< 1.05e2	n/a	n/a	n/a	104.0
S981002712	F	Cadmium - ICP-Fusion	ug/g	102.4	< 5.00e-03	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Cerium - ICP-Fusion	ug/g	100.4	< 2.00e-02	< 4.17e+02	< 4.18e2	n/a	n/a	n/a	418.0
S981002712	F	Cobalt - ICP-Fusion	ug/g	102.4	< 1.00e-02	< 2.08e+02	< 2.09e2	n/a	n/a	n/a	208.0
S981002712	F	Copper - ICP-Fusion	ug/g	102.8	< 1.00e-02	< 1.04e+03	< 1.05e3	n/a	n/a	n/a	1.04e+03
S981002712	F	Lanthanum - ICP-Fusion	ug/g	102.8	< 1.00e-02	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Magnesium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 1.04e+03	< 1.05e3	n/a	n/a	n/a	1.04e+03
S981002712	F	Molybdenum - ICP-Fusion	ug/g	100.2	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Niobium - ICP-Fusion	ug/g	102.2	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Phosphorus - ICP-Fusion	ug/g	103.2	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Lead - ICP-Fusion	ug/g	98.60	< 1.00e-01	< 1.10e+04	< 7.58e+03	2.37e+04	138	n/a	4.18e+03
S981002712	F	Sulfur - ICP-Fusion	ug/g	102.8	< 1.00e-02	< 1.25e+03	< 1.26e3	9.29e+03	36.8	n/a	1.25e+03
S981002712	F	Antimony - ICP-Fusion	ug/g	98.60	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Selenium - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.08e+03	< 2.09e3	n/a	n/a	n/a	2.08e+03
S981002712	F	Strontium - ICP-Fusion	ug/g	100.0	< 1.00e-02	< 2.08e+02	< 2.09e2	n/a	n/a	n/a	208.0
S981002712	F	Titanium - ICP-Fusion	ug/g	103.2	< 1.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	4.18e+03
S981002712	F	Thallium - ICP-Fusion	ug/g	98.60	< 1.00e-02	< 1.04e+03	< 1.05e3	n/a	n/a	n/a	1.04e+03
S981002712	F	Vanadium - ICP-Fusion	ug/g	101.0	< 1.00e-02	< 4.18e+02	< 4.19e2	n/a	n/a	n/a	4.18e+02
S981002712	F	Vanadium - ICP-Fusion	ug/g	104.3	< 1.00e-02	< 4.18e+02	< 4.19e2	n/a	n/a	n/a	4.18e+02
S981002712	F	Cobalt-60 by GEA	uCi/g	13.5	< 8.00e-05	< 3.982	< 5.87e0	n/a	n/a	n/a	5.982
S981002713	A	Sn117 ICP/MS ActDig LA-505159	ug/g	99.00	< 2.00e-02	< 4.17e+03	< 4.18e3	3.192	9.70	n/a	8.00e-02
S981002713	A	Cs133 ICP/MS ActDig LA-505159	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	4.845	15.1	n/a	8.00e-02
S981002713	A	Cs135 ICP/MS ActDig LA-505159	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	10.01	13.7	n/a	8.00e-02
S981002713	A	Cs137 ICP/MS ActDig LA-505159	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Protactinium-231 by ICP/MS	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Thorium-229 by ICP/MS	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Thorium-230 by ICP/MS	ug/g	n/a	< 2.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Thorium-232 by ICP/MS	ug/g	107.0	< 1.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Pu241 by ICP/MS	ug/g	103.9	< 1.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Pu244/Am244 by ICP/MS	ug/g	n/a	< 1.00e-02	< 4.17e+03	< 4.18e3	n/a	n/a	n/a	8.00e-02
S981002713	A	Arsenic - ICP-Acid Digest	ug/g	86.30	< 1.00e-02	< 18.00	< 15.10	16.55	17.5	n/a	9.900
S981002713	A	Beryllium - ICP-Acid Digest	ug/g	95.60	< 1.00e-01	< 99.20	< 97.20e1	n/a	n/a	n/a	99.00
S981002713	A	Cadmium - ICP-Acid Digest	ug/g	102.2	< 5.00e-03	< 4.960	< 4.86e0	n/a	n/a	n/a	4.960
S981002713	A	Cerium - ICP-Acid Digest	ug/g	95.20	< 5.00e-03	< 11.50	< 12.70	12.10	9.92	n/a	4.960
S981002713	A	Cobalt - ICP-Acid Digest	ug/g	97.20	< 1.00e-01	< 99.20	< 97.20e1	n/a	n/a	n/a	99.00
S981002713	A	Copper - ICP-Acid Digest	ug/g	94.00	< 2.00e-02	< 9.920	< 11.95e1	n/a	n/a	n/a	9.900
S981002713	A	Lanthanum - ICP-Acid Digest	ug/g	88.60	< 1.00e-02	< 4.960	< 4.86e0	n/a	n/a	n/a	4.960
S981002713	A	Magnesium - ICP-Acid Digest	ug/g	97.00	< 5.00e-02	< 49.60	< 47.8e0	n/a	n/a	n/a	99.00
S981002713	A	Molybdenum - ICP-Acid Digest	ug/g	99.00	< 1.00e-01	< 9.610	< 9.08e0	934.5	5.67	n/a	99.00
S981002713	A	Niobium - ICP-Acid Digest	ug/g	95.60	< 5.00e-02	< 9.920	< 4.86e1	n/a	n/a	n/a	49.60
S981002713	A	Neodymium - ICP-Acid Digest	ug/g	96.00	< 1.00e-01	< 99.20	< 97.20e1	n/a	n/a	n/a	99.00

5058

5058

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002713	A	Phosphorus -ICP-Acid Digest	µg/g	99.60	<2.00e-01	1.92e-01	2.77e-01	2.32e-01	33.6	N/A	198.0
S981002713	A	Lead -ICP-Acid Digest	µg/g	99.60	<1.00e-01	3.72e-01	3.92e-01	3.84e-01	19.2	N/A	99.00
S981002713	A	Sulfur -ICP-Acid Digest	µg/g	99.60	<1.50e-01	3.69e-02	3.02e-02	3.34e-02	N/A	N/A	99.00
S981002713	A	Antimony -ICP-Acid Digest	µg/g	97.80	<6.00e-02	8.76e-03	8.92e-03	8.86e-03	2.14	N/A	59.50
S981002713	A	Selenium -ICP-Acid Digest	µg/g	97.80	<6.00e-02	<5.84e-03	<5.84e-03	N/A	N/A	N/A	59.00
S981002713	A	Samarium -ICP-Acid Digest	µg/g	97.80	<1.00e-01	<99.20	<9.79e-01	N/A	N/A	N/A	99.00
S981002713	A	Strontium -ICP-Acid Digest	µg/g	99.60	<1.00e-01	<99.20	<9.79e-01	N/A	N/A	N/A	99.00
S981002713	A	Titanium -ICP-Acid Digest	µg/g	99.60	<1.00e-02	9.79e-02	1.23e-02	1.10e-02	23.6	N/A	9.900
S981002713	A	Vanadium -ICP-Acid Digest	µg/g	99.60	<1.00e-02	1.58e-02	1.13e-02	1.35e-02	33.2	N/A	9.900
S981002713	A	Thallium -ICP-Acid Digest	µg/g	99.60	<2.00e-02	<1.98e-02	<1.95e-02	N/A	N/A	N/A	198.0
S981002713	A	Acetate by IC-DIONE	µg/g	99.60	<5.00e-02	<4.98e-02	<4.95e-02	N/A	N/A	N/A	49.60
S981002714	W	Glycolate IC-DIONE	µg/g	75.43	<2.00e-02	8.03e-02	4.88e-02	7.45e-02	15.9	N/A	82.36
S981002714	W	Glycolate IC-DIONE	µg/g	99.10	<5.00e-02	<2.31e-02	<2.28e-02	N/A	N/A	N/A	230.6

Appendix A. Opportunistic Analyte Results.
81-105 (2)CODE NUMBER: 246
SEGMENT #: 9

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count	Err%
S981002602	F	Silver - ICP-Fusion	ug/g	100.2	< 1.00e-02	< 1.97e+02	< 1.96e2	n/a	n/a	91.70	197.0	n/a	n/a
S981002602	F	Arsenic - ICP-Fusion	ug/g	104.2	< 1.00e-03	< 1.97e+03	< 1.96e3	n/a	n/a	105.0	1.97e+03	n/a	n/a
S981002602	F	Beryllium - ICP-Fusion	ug/g	103.2	< 1.00e-03	< 98.50	< 97.8e1	n/a	n/a	101.0	98.40	n/a	n/a
S981002602	F	Cadmium - ICP-Fusion	ug/g	102.2	< 1.00e-01	< 9.7e+03	< 9.7e3	n/a	n/a	101.0	1.97e+03	n/a	n/a
S981002602	F	Cerium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 3.94e+02	< 3.91e2	n/a	n/a	99.30	394.0	n/a	n/a
S981002602	F	Cobalt - ICP-Fusion	ug/g	96.80	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	94.80	984.0	n/a	n/a
S981002602	F	Copper - ICP-Fusion	ug/g	102.2	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Lanthanum - ICP-Fusion	ug/g	102.2	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Magnesium - ICP-Fusion	ug/g	103.4	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Molybdenum - ICP-Fusion	ug/g	103.4	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Neodymium - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Phosphorus - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Lead - ICP-Fusion	ug/g	97.60	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Sulfur - ICP-Fusion	ug/g	97.60	< 1.00e-01	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Antimony - ICP-Fusion	ug/g	98.14	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Selenium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Samarium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Strontium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Titanium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Vanadium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	101.0	984.0	n/a	n/a
S981002602	F	Thallium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 9.85e+02	< 9.78e2	n/a	n/a	108.0	1.97e+03	n/a	n/a
S981002602	F	Cobalt-60 by GEA	uCi/g	115.3	< 1.00e-03	< 46.02e+02	< 45.89e2	n/a	n/a	99.00	9.90	n/a	n/a
S981002602	F	Sn117 ICP/MS Acid Dig LA-505159	ug/g	99.00	< 1.00e-05	< 29.31	< 26.70	n/a	n/a	28.00	9.29	n/a	n/a
S981002602	F	Cs133 ICP/MS Acid Dig LA-505159	ug/g	n/a	< 5.00e-05	< 63.04	< 55.00	n/a	n/a	59.02	13.6	n/a	n/a
S981002602	F	Cs135 ICP/MS Acid Dig LA-505159	ug/g	n/a	< 5.00e-05	< 63.04	< 55.00	n/a	n/a	59.02	13.6	n/a	n/a
S981002602	F	Cs137 ICP/MS Acid Dig LA-505159	ug/g	n/a	< 5.00e-05	< 63.04	< 55.00	n/a	n/a	59.02	13.6	n/a	n/a
S981002602	F	Protactinium-231 by ICP/MS	ug/g	n/a	< 5.00e-05	< 22.36	< 21.9e1	n/a	n/a	n/a	n/a	n/a	n/a
S981002602	F	Thorium-229 by ICP/MS	ug/g	n/a	< 5.00e-05	< 22.36	< 21.9e1	n/a	n/a	n/a	n/a	n/a	n/a
S981002602	F	Thorium-230 by ICP/MS	ug/g	n/a	< 5.00e-05	< 22.36	< 21.9e1	n/a	n/a	n/a	n/a	n/a	n/a
S981002602	F	Thorium-232 by ICP/MS	ug/g	102.2	< 1.93e-05	< 7.750	< 7.58e0	n/a	n/a	99.00	107.2	n/a	n/a
S981002602	F	Pu/Am-241 by ICP/MS	ug/g	110.1	< 3.14e-05	< 15.40	< 14.4e1	n/a	n/a	n/a	n/a	n/a	n/a
S981002602	F	Am-241/Am-243 by ICP/MS	ug/g	n/a	< 3.14e-05	< 13.71	< 13.4e1	n/a	n/a	n/a	n/a	n/a	n/a
S981002602	F	Pu 244/Am 244 by ICP/MS	ug/g	82.00	< 1.00e-02	< 39.60	< 38.7e1	n/a	n/a	82.60	39.60	n/a	n/a
S981002602	F	Arsenic - ICP-Acid Digest	ug/g	90.60	< 1.00e-03	< 1.980	< 1.92e0	n/a	n/a	86.20	1.980	n/a	n/a
S981002602	F	Beryllium - ICP-Acid Digest	ug/g	101.8	< 1.00e-03	< 1.980	< 1.92e0	n/a	n/a	79.60	1.980	n/a	n/a
S981002602	F	Cadmium - ICP-Acid Digest	ug/g	88.40	< 1.00e-03	< 1.980	< 1.92e0	n/a	n/a	81.20	39.60	n/a	n/a
S981002602	F	Cerium - ICP-Acid Digest	ug/g	97.60	< 1.00e-01	< 39.60	< 38.7e1	n/a	n/a	69.00	7.920	n/a	n/a
S981002602	F	Cobalt - ICP-Acid Digest	ug/g	89.20	< 1.00e-02	< 3.960	< 3.87e0	n/a	n/a	84.80	39.60	n/a	n/a
S981002602	F	Copper - ICP-Acid Digest	ug/g	90.20	< 1.00e-02	< 19.80	< 19.2e0	n/a	n/a	84.80	39.60	n/a	n/a
S981002602	F	Lanthanum - ICP-Acid Digest	ug/g	92.60	< 1.00e-01	< 1.37e+03	< 1.23e+03	n/a	n/a	81.20	39.60	n/a	n/a
S981002602	F	Magnesium - ICP-Acid Digest	ug/g	92.60	< 1.00e-01	< 1.37e+03	< 1.23e+03	n/a	n/a	81.20	39.60	n/a	n/a
S981002602	F	Molybdenum - ICP-Acid Digest	ug/g	92.60	< 1.00e-02	< 19.80	< 19.2e0	n/a	n/a	80.60	39.60	n/a	n/a
S981002602	F	Neodymium - ICP-Acid Digest	ug/g	97.60	< 1.00e-01	< 1.51e+02	< 130.0	n/a	n/a	81.20	39.60	n/a	n/a

Sample#	R#	Analyle	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002608	A	Phosphorus - ICP-Acid Digest	ug/g	90.60	<2.00e-01	1.77e+04	1.53e+04	1.62e+04	11.7	-2.50e1	79.20	n/a
S98T002608	A	Lead - ICP-Acid Digest	ug/g	85.20	<1.00e-01	6.35e+02	584.0	609.0	8.21	17.00	39.20	n/a
S98T002608	A	Sulfur - ICP-Acid Digest	ug/g	84.40	<1.00e-01	1.01e+03	881.0	945.3	13.6	186.0	39.20	n/a
S98T002608	A	Antimony - ICP-Acid Digest	ug/g	88.60	<6.00e-02	< 23.70	<2.32e1	n/a	n/a	79.20	39.20	n/a
S98T002608	A	Selenium - ICP-Acid Digest	ug/g	88.60	<1.00e-01	< 39.60	<3.87e1	n/a	n/a	79.20	39.20	n/a
S98T002608	A	Samarium - ICP-Acid Digest	ug/g	97.00	<1.00e-01	< 39.60	<3.87e1	n/a	n/a	82.40	39.20	n/a
S98T002608	A	Strontium - ICP-Acid Digest	ug/g	97.00	<1.00e-02	2.28e+04	1.97e+04	2.09e+04	10.8	92.40	3.960	n/a
S98T002608	A	Titanium - ICP-Acid Digest	ug/g	96.20	<1.00e-02	1.23e+02	<1.05e0	115.0	17.4	82.30	3.960	n/a
S98T002608	A	Thallium - ICP-Acid Digest	ug/g	86.80	<2.00e-01	< 19.10	<1.94e1	n/a	n/a	79.20	79.20	n/a
S98T002608	A	Vanadium - ICP-Acid Digest	ug/g	90.80	<3.00e-02	< 19.20	<1.96e1	n/a	n/a	78.80	19.80	n/a
S98T002614	W	Acetate by IC-BIOMEX 4000/4500	ug/g	93.97	2.90e-02	1.41e+03	1.49e+03	1.45e+03	5.52	119.0	33.42	n/a
S98T002614	W	Glycolate IC-BIOMEX 4000/4500	ug/g	93.32	<3.60e-02	< 65.57	<6.76e1	n/a	n/a	100.0	65.56	n/a

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 246
SEGMENT #: 9A

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
5981002603	F	Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.05e+02	<2.09e2	N/A	N/A	N/A	205.0	N/A	N/A
5981002603	F	Arsenic - ICP-Fusion	ug/g	104.2	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Beryllium - ICP-Fusion	ug/g	103.2	<5.00e-03	<1.03e+02	<1.04e2	N/A	N/A	N/A	102.0	N/A	N/A
5981002603	F	Cadmium - ICP-Fusion	ug/g	102.2	<5.00e-03	<1.03e+02	<1.04e2	N/A	N/A	N/A	102.0	N/A	N/A
5981002603	F	Cerium - ICP-Fusion	ug/g	102.8	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Cobalt - ICP-Fusion	ug/g	100.2	<1.00e-02	<4.10e+02	<4.17e2	N/A	N/A	N/A	410.0	N/A	N/A
5981002603	F	Copper - ICP-Fusion	ug/g	96.80	<1.00e-02	<2.05e+02	<2.09e2	N/A	N/A	N/A	205.0	N/A	N/A
5981002603	F	Lanthanum - ICP-Fusion	ug/g	102.2	<5.00e-02	<1.03e+03	<1.04e3	N/A	N/A	N/A	1.02e+03	N/A	N/A
5981002603	F	Magnesium - ICP-Fusion	ug/g	102.6	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Molybdenum - ICP-Fusion	ug/g	101.4	<5.00e-02	<1.03e+03	<1.04e3	N/A	N/A	N/A	1.02e+03	N/A	N/A
5981002603	F	Neodymium - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Phosphorus - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.38e+04	2.19e+04	2.28e+04	8.32	N/A	4.10e+03	N/A	N/A
5981002603	F	Lead - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Sulfur - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Tantalum - ICP-Fusion	ug/g	96.00	<1.00e-01	<2.38e+03	<2.09e3	N/A	N/A	N/A	1.23e+03	N/A	N/A
5981002603	F	Samarium - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.05e+03	<2.09e3	N/A	N/A	N/A	2.05e+03	N/A	N/A
5981002603	F	Strontium - ICP-Fusion	ug/g	102.6	<1.00e-02	<2.05e+04	2.51e+04	2.48e+04	2.01	N/A	205.0	N/A	N/A
5981002603	F	Titanium - ICP-Fusion	ug/g	103.0	<1.00e-02	<2.05e+02	<2.09e2	N/A	N/A	N/A	205.0	N/A	N/A
5981002603	F	Thallium - ICP-Fusion	ug/g	96.60	<2.00e-01	<4.10e+03	<4.17e3	N/A	N/A	N/A	4.10e+03	N/A	N/A
5981002603	F	Vanadium - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.03e+03	<1.04e3	N/A	N/A	N/A	1.02e+03	N/A	N/A
5981002603	F	Cobalt-60 by GEA	uCi/g	108.7	<2.99e-05	<5.690	<6.44e0	N/A	N/A	N/A	5.690	N/A	N/A
5981002609	A	Sn117 ICP/MS ActDtg LA-505159	ug/g	115.3	<3.00e-03	<4.19e+02	<5.82e2	N/A	N/A	N/A	619.4	N/A	N/A
5981002609	A	Cs133 ICP/MS ActDtg LA-505159	ug/g	99.00	<2.00e-05	<8.259	<7.76e0	N/A	N/A	N/A	8.258	N/A	N/A
5981002609	A	Cs137 ICP/MS ActDtg LA-505159	ug/g	N/A	<5.00e-05	<34.28	<29.10	31.69	16.4	N/A	8.258	N/A	N/A
5981002609	A	Protectinium-231 by ICP/MS	ug/g	N/A	<8.50e-05	<67.75	<60.20	63.97	11.7	N/A	8.258	N/A	N/A
5981002609	A	Thorium-229 by ICP/MS	ug/g	N/A	<5.57e-05	<23.00	<2.16e1	N/A	N/A	N/A	23.00	N/A	N/A
5981002609	A	Thorium-230 by ICP/MS	ug/g	N/A	<4.46e-05	<18.43	<1.75e1	N/A	N/A	N/A	18.43	N/A	N/A
5981002609	A	Thorium-232 by ICP/MS	ug/g	N/A	<3.55e-05	<14.67	<1.38e1	N/A	N/A	N/A	14.67	N/A	N/A
5981002609	A	Pu/Am-241 by ICP/MS	ug/g	102.2	<1.91e-05	<7.971	<7.49e0	N/A	N/A	N/A	7.970	N/A	N/A
5981002609	A	Pu/Am-243 by ICP/MS	ug/g	110.1	<3.71e-05	<15.30	<1.44e1	N/A	N/A	N/A	15.30	N/A	N/A
5981002609	A	Am 243 Cm 244 by ICP/MS	ug/g	N/A	<8.82e-05	<15.83	<1.49e1	N/A	N/A	N/A	15.83	N/A	N/A
5981002609	A	Pu 244 Cm 244 by ICP/MS	ug/g	N/A	<3.41e-05	<14.10	<1.33e1	N/A	N/A	N/A	14.10	N/A	N/A
5981002609	A	Silver - ICP-Acid Digest	ug/g	82.00	<1.00e-02	<24.60	<19.60	22.10	22.6	N/A	4.060	N/A	N/A
5981002609	A	Arsenic - ICP-Acid Digest	ug/g	90.60	<5.00e-01	<40.70	<3.82e1	N/A	N/A	N/A	40.60	N/A	N/A
5981002609	A	Beryllium - ICP-Acid Digest	ug/g	101.8	<5.00e-03	<3.600	<3.62e0	N/A	N/A	N/A	2.060	N/A	N/A
5981002609	A	Cadmium - ICP-Acid Digest	ug/g	88.60	<5.00e-03	<4.360	<4.91e0	N/A	N/A	N/A	2.060	N/A	N/A
5981002609	A	Cerium - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<4.700	<3.82e0	N/A	N/A	N/A	40.60	N/A	N/A
5981002609	A	Cobalt - ICP-Acid Digest	ug/g	89.20	<2.00e-02	<22.20	<16.30	19.22	30.6	N/A	8.160	N/A	N/A
5981002609	A	Copper - ICP-Acid Digest	ug/g	90.20	<1.00e-02	<4.070	<3.82e0	N/A	N/A	N/A	40.60	N/A	N/A
5981002609	A	Lanthanum - ICP-Acid Digest	ug/g	98.00	<5.00e-02	<25.20	<1.91e0	N/A	N/A	N/A	20.40	N/A	N/A
5981002609	A	Magnesium - ICP-Acid Digest	ug/g	93.80	<1.00e-01	<1.43e+03	<1.28e3	1.36e+03	11.1	N/A	40.60	N/A	N/A
5981002609	A	Molybdenum - ICP-Acid Digest	ug/g	92.40	<5.00e-02	<20.30	<17.9e0	N/A	N/A	N/A	20.40	N/A	N/A
5981002609	A	Neodymium - ICP-Acid Digest	ug/g	97.60	<1.00e-01	<1.75e+02	<153.10	163.0	12.3	N/A	40.60	N/A	N/A

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002609	A	Phosphorus - CP-Acid Digest	ug/g	90.60	<2.00e-01	2.11e+04	2.31e+04	2.21e+04	9.05	n/a	81.40	n/a
S981002609	A	Lead - CP-Acid Digest	ug/g	85.20	<1.00e-01	8.40e+02	758.0	799.0	10.3	n/a	40.60	n/a
S981002609	A	Sulfur - CP-Acid Digest	ug/g	84.40	<1.00e-01	3.26e+03	805.0	2.03e+03	121	n/a	40.60	n/a
S981002609	A	Antimony - CP-Acid Digest	ug/g	88.60	<6.00e-02	<24.40	<2.29e1	n/a	n/a	n/a	24.40	n/a
S981002609	A	Selenium - CP-Acid Digest	ug/g	88.60	<1.00e-01	<40.70	<3.82e1	n/a	n/a	n/a	40.60	n/a
S981002609	A	Strontium - CP-Acid Digest	ug/g	97.40	<1.00e-01	<40.70	<3.82e1	n/a	n/a	n/a	40.60	n/a
S981002609	A	Zinc - CP-Acid Digest	ug/g	97.00	<1.00e-02	2.52e+04	2.27e+04	2.40e+04	10.4	n/a	40.60	n/a
S981002609	A	Thallium - CP-Acid Digest	ug/g	86.80	<2.00e-01	<81.40	<7.65e1	75.55	14.2	n/a	4.060	n/a
S981002609	A	Vanadium - CP-Acid Digest	ug/g	90.80	<5.00e-02	<20.30	<1.91e1	n/a	n/a	n/a	81.40	n/a
S981002615	W	Acetate by IC-DIONEX 4000/4500	ug/g	93.97	2.90e-02	1.89e+03	1.90e+03	1.90e+03	0.53	n/a	24.02	n/a
S981002615	W	Glycolate IC-DIONEX 4000/4500	ug/g	93.52	<5.60e-02	<67.27	<6.53e1	n/a	n/a	n/a	67.26	n/a

Appendix A. Opportunistic Analyte Results.
BR-105 (2)

CORE NUMBER: 245

SEGMENT #: Core Composite

SEGMENT PORTION: Core Composite

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk	Det Limit	Count	Err%
S98T003392		F Silver - ICP-Fusion	ug/g	98.60	< 1.00e-02	< 2.04e-02	< 2.022	n/a	n/a	92.10	204.0	n/a
S98T003392		F Arsenic - ICP-Fusion	ug/g	100.6	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	101.0	2.04e+03	n/a
S98T003392		F Barium - ICP-Fusion	ug/g	101.4	< 5.00e-03	< 1.02e-02	< 1.012	n/a	n/a	98.80	102.0	n/a
S98T003392		F Beryllium - ICP-Fusion	ug/g	101.8	< 5.00e-03	< 1.02e-02	< 1.012	n/a	n/a	101.0	102.0	n/a
S98T003392		F Cadmium - ICP-Fusion	ug/g	100.4	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	99.10	2.04e+03	n/a
S98T003392		F Cobalt - ICP-Fusion	ug/g	100.4	< 2.00e-02	< 4.07e-02	< 4.042	n/a	n/a	99.30	407.0	n/a
S98T003392		F Copper - ICP-Fusion	ug/g	95.60	< 1.00e-02	< 2.04e-02	< 2.022	n/a	n/a	93.60	204.0	n/a
S98T003392		F Lanthanum - ICP-Fusion	ug/g	100.4	< 5.00e-02	< 1.02e-03	< 1.013	n/a	n/a	97.90	1.02e+03	n/a
S98T003392		F Magnesium - ICP-Fusion	ug/g	103.6	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	101.0	2.04e+03	n/a
S98T003392		F Molybdenum - ICP-Fusion	ug/g	105.4	< 5.00e-02	< 1.02e-03	< 1.013	n/a	n/a	105.0	1.02e+03	n/a
S98T003392		F Neodymium - ICP-Fusion	ug/g	97.60	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	94.90	2.04e+03	n/a
S98T003392		F Phosphorus - ICP-Fusion	ug/g	105.2	< 2.00e-01	< 4.07e-03	< 4.043	n/a	n/a	99.80	4.07e+03	n/a
S98T003392		F Lead - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	103.0	2.04e+03	n/a
S98T003392		F Selenium - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	106.0	2.04e+03	n/a
S98T003392		F Strontium - ICP-Fusion	ug/g	96.40	< 1.00e-01	< 2.04e-03	< 2.023	n/a	n/a	98.50	2.04e+03	n/a
S98T003392		F Thallium - ICP-Fusion	ug/g	97.80	< 1.00e-02	< 2.04e-02	< 2.022	n/a	n/a	95.20	204.0	n/a
S98T003392		F Vanadium - ICP-Fusion	ug/g	95.20	< 2.00e-01	< 4.07e-03	< 4.043	n/a	n/a	92.30	4.07e+03	n/a
S98T003392		F Vanadium - ICP-Fusion	ug/g	100.2	< 5.00e-02	< 1.02e-03	< 1.013	n/a	n/a	99.20	1.02e+03	n/a
S98T003392		F Cobalt-60 by GEA	uCi/g	109.8	< 7.49e-03	< 7.86e-3	< 7.86e-3	n/a	n/a	n/a	7.00e-03	n/a
S98T003394		A Sm171 ICP/MS Acid Dig LA-505159	ug/g	115.3	4.00e-03	< 5.926	< 6.2060	n/a	n/a	n/a	5.925	n/a
S98T003394		A Cs133 ICP/MS Acid Dig LA-505159	ug/g	99.00	< 2.00e-03	2.84e-01	3.35e-01	3.10e-01	16.5	n/a	7.90e-02	n/a
S98T003394		A Cs135 ICP/MS Acid Dig LA-505159	ug/g	n/a	4.30e-05	1.38e-01	1.24e-01	1.31e-01	10.7	n/a	7.90e-02	n/a
S98T003394		A Cs137 ICP/MS Acid Dig LA-505159	ug/g	n/a	6.80e-05	4.43e-01	2.68e-01	3.55e-01	49.2	n/a	7.90e-02	n/a
S98T003394		A Protactinium-231 by ICP/MS	ug/g	n/a	< 5.57e-05	< 2.30e-01	< 2.30e-1	n/a	n/a	n/a	1.76e-01	n/a
S98T003394		A Thorium-229 by ICP/MS	ug/g	n/a	< 4.46e-05	< 1.76e-01	< 1.84e-1	n/a	n/a	n/a	1.40e-01	n/a
S98T003394		A Thorium-230 by ICP/MS	ug/g	n/a	< 3.55e-05	< 1.40e-01	< 1.47e-1	n/a	n/a	n/a	1.60e-01	n/a
S98T003394		A Thorium-232 by ICP/MS	ug/g	103.6	2.12e-05	1.093	1.220	1.156	11.3	99.50	7.60e-02	n/a
S98T003394		A Pu/Am-241 by ICP/MS	ug/g	107.8	< 3.71e-05	< 1.46e-01	< 1.53e-1	n/a	n/a	106.2	1.46e-01	n/a
S98T003394		A Pu/Am-243 by ICP/MS	ug/g	n/a	< 8.85e-05	< 1.51e-01	< 1.47e-1	n/a	n/a	n/a	1.51e-01	n/a
S98T003394		A Pu 244/Am 244 by ICP/MS	ug/g	n/a	< 3.41e-05	< 1.35e-01	< 1.41e-1	n/a	n/a	n/a	1.35e-01	n/a
S98T003394		A Silver - ICP-Acid Digest	ug/g	85.00	< 1.00e-02	< 17.90	< 17.00	17.45	5.16	95.00	9.850	n/a
S98T003394		A Arsenic - ICP-Acid Digest	ug/g	92.20	< 1.00e-02	< 98.30	< 98.30	< 1.0362	n/a	101.4	98.50	n/a
S98T003394		A Beryllium - ICP-Acid Digest	ug/g	103.0	< 5.00e-03	< 4.910	< 5.1460	n/a	n/a	103.2	4.920	n/a
S98T003394		A Cadmium - ICP-Acid Digest	ug/g	93.00	< 5.00e-03	< 4.910	< 5.1460	n/a	n/a	103.6	4.920	n/a
S98T003394		A Cerium - ICP-Acid Digest	ug/g	97.20	< 1.00e-01	< 98.30	< 1.0362	n/a	n/a	100.6	98.50	n/a
S98T003394		A Cobalt - ICP-Acid Digest	ug/g	93.40	< 2.00e-02	< 19.70	< 2.0661	n/a	n/a	101.8	19.70	n/a
S98T003394		A Copper - ICP-Acid Digest	ug/g	88.40	< 1.00e-02	< 9.830	< 1.0361	n/a	n/a	93.60	9.850	n/a
S98T003394		A Lanthanum - ICP-Acid Digest	ug/g	97.00	< 5.00e-02	< 49.10	< 5.1461	n/a	n/a	100.0	49.10	n/a
S98T003394		A Magnesium - ICP-Acid Digest	ug/g	94.00	< 1.00e-01	< 98.30	< 1.0362	n/a	n/a	105.0	98.50	n/a
S98T003394		A Molybdenum - ICP-Acid Digest	ug/g	95.00	< 5.00e-02	< 49.10	< 5.1461	n/a	n/a	102.6	49.10	n/a
S98T003394		A Neodymium - ICP-Acid Digest	ug/g	95.80	< 1.00e-01	< 98.30	< 1.0362	n/a	n/a	98.40	98.50	n/a

Sample#	R. #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981003394	A	Phosphorus - ICP-Acid Digest	ug/g	91.20	<2.00e-01	4.95e+02	372.0	552.0	15.0	101.5	195.0	N/A
S981003394	A	Lead - ICP-Acid Digest	ug/g	88.80	<1.00e-01	96.30	110.92	97.0	N/A	102.4	98.50	N/A
S981003394	A	Sulfur - ICP-Acid Digest	ug/g	90.40	1.31e-01	5.21e+02	579.0	557.3	20.4	101.0	98.50	N/A
S981003394	A	Antimony - ICP-Acid Digest	ug/g	90.40	<6.00e-02	37.40	46.76	N/A	N/A	103.5	98.50	N/A
S981003394	A	Selenium - ICP-Acid Digest	ug/g	91.50	<1.00e-01	98.30	110.92	N/A	N/A	107.2	98.50	N/A
S981003394	A	Strontium - ICP-Acid Digest	ug/g	92.60	<1.00e-02	9.830	110.92	N/A	N/A	97.20	98.50	N/A
S981003394	A	Strontium - ICP-Acid Digest	ug/g	92.60	<1.00e-02	9.830	110.92	N/A	N/A	97.20	98.50	N/A
S981003394	A	Thallium - ICP-Acid Digest	ug/g	93.60	<2.00e-01	1.97e+03	2.08e+04	N/A	N/A	101.8	9.850	N/A
S981003394	A	Vanadium - ICP-Acid Digest	ug/g	83.40	<5.00e-02	2.08e+02	15.14	N/A	N/A	99.40	195.0	N/A
S981003394	A	Vanadium - ICP-Acid Digest	ug/g	83.40	<5.00e-02	2.08e+02	15.14	N/A	N/A	99.40	195.0	N/A
S981003395	W	Glycolate IC-DINEX 4000/4500	ug/g	92.71	<5.00e-02	66.98	52.20	59.59	24.8	91.84	44.52	N/A
S981003395	W	Glycolate IC-DINEX 4000/4500	ug/g	92.71	<5.00e-02	66.98	52.20	59.59	24.8	91.84	44.52	N/A
S981003398	F	Silver - ICP-Fusion	ug/g	98.60	<1.00e-02	2.08e+03	2.07e3	N/A	N/A	91.90	207.0	N/A
S981003398	F	Arsenic - ICP-Fusion	ug/g	100.6	<1.00e-01	2.08e+03	2.07e3	N/A	N/A	100.0	2.07e+03	N/A
S981003398	F	Beryllium - ICP-Fusion	ug/g	101.4	<5.00e-03	1.03e+02	1.03e2	N/A	N/A	97.60	103.0	N/A
S981003398	F	Cadmium - ICP-Fusion	ug/g	101.4	<5.00e-03	1.03e+02	1.03e2	N/A	N/A	97.60	103.0	N/A
S981003398	F	Cerium - ICP-Fusion	ug/g	100.4	<1.00e-02	2.08e+03	4.13e2	N/A	N/A	97.80	2.07e+03	N/A
S981003398	F	Cobalt - ICP-Fusion	ug/g	100.4	<1.00e-02	2.08e+03	4.13e2	N/A	N/A	100.0	4.14e2	N/A
S981003398	F	Copper - ICP-Fusion	ug/g	95.60	<1.00e-02	2.08e+03	1.03e3	N/A	N/A	91.20	207.0	N/A
S981003398	F	Lanthanum - ICP-Fusion	ug/g	100.4	<5.00e-02	2.08e+03	1.03e3	N/A	N/A	96.20	1.03e+03	N/A
S981003398	F	Magnesium - ICP-Fusion	ug/g	103.6	<1.00e-02	2.08e+03	1.03e3	N/A	N/A	107.0	2.07e+03	N/A
S981003398	F	Molybdenum - ICP-Fusion	ug/g	105.4	<5.00e-02	2.08e+03	1.03e3	N/A	N/A	105.0	1.03e+03	N/A
S981003398	F	Neodymium - ICP-Fusion	ug/g	97.60	<1.00e-01	2.08e+03	1.03e3	N/A	N/A	94.30	2.07e+03	N/A
S981003398	F	Phosphorus - ICP-Fusion	ug/g	105.2	<2.00e-01	2.19e+04	2.08e+04	2.14e+04	5.15	103.0	4.14e+03	N/A
S981003398	F	Lead - ICP-Fusion	ug/g	100.0	<1.00e-01	2.08e+03	2.07e3	N/A	N/A	105.0	2.07e+03	N/A
S981003398	F	Sulfur - ICP-Fusion	ug/g	102.4	<1.00e-01	2.08e+03	2.07e3	N/A	N/A	110.0	2.07e+03	N/A
S981003398	F	Antimony - ICP-Fusion	ug/g	106.0	<6.00e-02	2.08e+03	2.07e3	N/A	N/A	107.0	1.24e+03	N/A
S981003398	F	Selenium - ICP-Fusion	ug/g	96.40	<1.00e-01	2.08e+03	2.07e3	N/A	N/A	97.30	2.07e+03	N/A
S981003398	F	Samarium - ICP-Fusion	ug/g	97.80	<1.00e-02	2.08e+03	2.07e3	N/A	N/A	93.20	2.07e+03	N/A
S981003398	F	Strontium - ICP-Fusion	ug/g	97.80	<1.00e-02	2.34e+04	2.28e+04	2.31e+04	2.60	92.30	207.0	N/A
S981003398	F	Titanium - ICP-Fusion	ug/g	106.2	<1.00e-02	2.08e+03	2.07e3	N/A	N/A	104.0	207.0	N/A
S981003398	F	Thallium - ICP-Fusion	ug/g	95.20	<1.00e-01	4.13e+03	4.13e3	N/A	N/A	91.90	4.14e+03	N/A
S981003398	F	Vanadium - ICP-Fusion	ug/g	100.2	<5.00e-02	1.03e+03	1.03e3	N/A	N/A	99.00	1.03e+03	N/A
S981003398	F	Cobalt-60 by GEA	uCi/g	109.8	<7.49e-03	8.00e-01	8.00e-01	N/A	N/A	N/A	8.00e-01	N/A
S981003399	A	Sn117 ICP/MS ActDigi LA-505159	ug/g	115.3	4.00e-03	<6.1e+02	<5.84e2	N/A	N/A	N/A	610.7	N/A
S981003399	A	Cs133 ICP/MS ActDigi LA-505159	ug/g	99.00	<2.00e-03	<8.143	<7.7960	N/A	N/A	N/A	8.143	N/A
S981003399	A	Cs135 ICP/MS ActDigi LA-505159	ug/g	N/A	4.30e-05	28.09	30.80	29.45	9.17	N/A	8.143	N/A
S981003399	A	Cs137 ICP/MS ActDigi LA-505159	ug/g	N/A	6.80e-05	56.60	62.70	59.65	10.2	N/A	8.143	N/A
S981003399	A	Protactinium-231 by ICP/MS	ug/g	N/A	<5.37e-05	<2.68	<2.17e1	N/A	N/A	N/A	22.68	N/A
S981003399	A	Thorium-229 by ICP/MS	ug/g	N/A	<4.46e-05	<18.17	<1.74e1	N/A	N/A	N/A	18.17	N/A
S981003399	A	Thorium-230 by ICP/MS	ug/g	N/A	<3.53e-05	<14.46	<1.38e1	N/A	N/A	N/A	14.46	N/A
S981003399	A	Thorium-232 by ICP/MS	ug/g	103.6	2.12e-05	<7.859	<7.5260	N/A	N/A	99.50	7.859	N/A
S981003399	A	Pu/Am-241 by ICP/MS	ug/g	107.8	<3.71e-05	<15.09	<1.49e1	N/A	N/A	106.7	15.09	N/A
S981003399	A	Am-243/Cm-243 by ICP/MS	ug/g	N/A	<3.83e-05	<15.61	<1.46e1	N/A	N/A	N/A	15.61	N/A
S981003399	A	Pu-244/Cm-244 by ICP/MS	ug/g	N/A	<3.47e-05	<13.90	<1.35e1	N/A	N/A	N/A	13.90	N/A
S981003399	A	Silver - ICP-Acid Digest	ug/g	85.00	<1.00e-02	17.00	17.10	17.05	0.59	85.00	4.020	N/A
S981003399	A	Arsenic - ICP-Acid Digest	ug/g	92.20	<1.00e-01	<40.10	<3.84e1	N/A	N/A	92.20	40.20	N/A
S981003399	A	Beryllium - ICP-Acid Digest	ug/g	103.0	<5.00e-03	<2.010	<1.92e0	N/A	N/A	94.40	2.000	N/A
S981003399	A	Cadmium - ICP-Acid Digest	ug/g	93.00	<5.00e-03	<2.010	<1.92e0	N/A	N/A	89.20	2.000	N/A
S981003399	A	Cerium - ICP-Acid Digest	ug/g	97.20	<1.00e-01	<40.10	<3.84e1	N/A	N/A	89.20	40.20	N/A
S981003399	A	Cobalt - ICP-Acid Digest	ug/g	93.40	<2.00e-02	16.90	18.00	17.45	6.30	84.40	8.020	N/A

Sample#	P. #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003399	A	Copper - ICP-Acid Digest	ug/g	88.40	<1.00e-02	< 6.010	<1.84e0	n/a	n/a	56.80	4.020	n/a
S98T003399	A	Lead - ICP-Acid Digest	ug/g	97.00	<5.00e-02	< 20.10	<1.92e+1	n/a	n/a	92.20	20.00	n/a
S98T003399	A	Mercury - ICP-Acid Digest	ug/g	94.00	<1.00e-01	1.25e+03	1.38e+03	1.32e+03	9.89	106.4	40.20	n/a
S98T003399	A	Molybdenum - ICP-Acid Digest	ug/g	95.00	<5.00e-02	< 20.10	<1.92e+1	n/a	n/a	90.20	20.00	n/a
S98T003399	A	Nickel - ICP-Acid Digest	ug/g	95.80	<1.00e-01	1.40e+02	155.0	147.5	10.2	90.60	40.20	n/a
S98T003399	A	Phosphorus - ICP-Acid Digest	ug/g	91.20	<2.00e-01	1.79e+02	2.04e+04	1.92e+04	13.1	384.0	80.20	n/a
S98T003399	A	Selenium - ICP-Acid Digest	ug/g	88.80	<1.00e-01	6.68e+02	748.0	708.0	11.3	98.20	40.20	n/a
S98T003399	A	Sulfur - ICP-Acid Digest	ug/g	88.20	1.31e-01	1.50e+03	965.0	1.23e+03	43.4	310.0	40.20	n/a
S98T003399	A	Antimony - ICP-Acid Digest	ug/g	90.40	<5.00e-02	< 24.10	<2.30e+1	n/a	n/a	91.00	24.00	n/a
S98T003399	A	Selenium - ICP-Acid Digest	ug/g	91.20	<1.00e-01	< 40.10	<3.84e+1	n/a	n/a	90.60	40.20	n/a
S98T003399	A	Strontium - ICP-Acid Digest	ug/g	95.20	<1.00e-01	< 40.10	<3.84e+1	n/a	n/a	89.60	40.20	n/a
S98T003399	A	Strontium - ICP-Acid Digest	ug/g	96.00	<1.00e-02	2.01e+04	2.26e+04	2.14e+04	11.7	444.0	4.020	n/a
S98T003399	A	Titanium - ICP-Acid Digest	ug/g	96.40	<1.00e-02	< 87.70	96.60	92.15	9.66	92.40	4.020	n/a
S98T003399	A	Vanadium - ICP-Acid Digest	ug/g	87.60	<2.00e-01	< 80.20	<7.67e+1	n/a	n/a	81.40	80.20	n/a
S98T003399	A	Vanadium - ICP-Acid Digest	ug/g	93.40	<5.00e-02	< 20.10	<1.92e+1	n/a	n/a	88.60	20.00	n/a
S98T003400	U	Acetate by IC-DIONEX 4000/4500	ug/g	91.94	<2.00e-02	1.49e+03	1.34e+03	1.41e+03	10.6	89.52	43.41	n/a
S98T003400	U	Glycolate IC-DIONEX 4000/4500	ug/g	100.0	<5.60e-02	<1.22e+02	<1.21e2	n/a	n/a	87.48	121.6	n/a

25-mar-1999 09:38:43
A-0002-1

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 246
SEGMENT #: FIELD BLANK

SEGMENT PORTION: Field Blank

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S981002391		TIC by Acid/Coulometry	ug/mL	101.0	1.00e-01	< 5.000	<5.00e+01	n/a	n/a	n/a	5.000	n/a	n/a
S981002391		D Sn 117 by ICP/MS Acid Addition	ug/mL	105.5	<1.50e-04	<1.65e-03	<1.65e-3	n/a	n/a	n/a	2.00e-03	n/a	n/a
S981002391		D Cs 133 by ICP/MS Acid Addition	ug/mL	89.00	<2.00e-06	<2.20e-05	<2.20e-5	n/a	n/a	n/a	2.20e-05	n/a	n/a
S981002391		D Cs 135 by ICP/MS Acid Addition	ug/mL	n/a	<2.00e-06	<2.20e-05	<2.20e-5	n/a	n/a	n/a	2.20e-05	n/a	n/a
S981002391		D Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	<2.00e-06	3.08e-05	2.86e-05	2.97e-05	7.41	n/a	2.20e-05	n/a	n/a
S981002391		D Protection-231 by ICP/MS	ug/mL	n/a	44.72e-06	44.76e-04	44.76e-4	n/a	n/a	n/a	4.76e-04	n/a	n/a
S981002391		D Thorium-230 by ICP/MS	ug/mL	n/a	53.69e-06	53.70e-04	53.70e-4	n/a	n/a	n/a	5.70e-04	n/a	n/a
S981002391		D Thorium-232 by ICP/MS	ug/mL	n/a	55.69e-06	55.75e-04	55.75e-4	n/a	n/a	n/a	5.75e-04	n/a	n/a
S981002391		D Pu-243 by ICP/MS	ug/mL	105.0	<5.45e-06	55.00e-04	55.00e-4	n/a	n/a	n/a	5.00e-04	n/a	n/a
S981002391		D Pu-244 by ICP/MS	ug/mL	112.6	<5.35e-06	55.41e-04	55.41e-4	n/a	n/a	n/a	5.41e-04	n/a	n/a
S981002391		D Pu-244 by ICP/MS	ug/mL	n/a	44.64e-06	44.69e-04	44.69e-4	n/a	n/a	n/a	4.69e-04	n/a	n/a
S981002391		D Si-29 by ICP/MS	ug/mL	98.80	<3.35e-06	53.39e-04	53.39e-4	n/a	n/a	n/a	3.39e-04	n/a	n/a
S981002391		D Si-29 by ICP/MS	ug/mL	102.2	<1.00e-01	<1.00e-02	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Arsenic-ICP-Acid Dill.	ug/mL	102.6	<5.00e-03	55.00e-03	55.00e-3	n/a	n/a	n/a	5.00e-03	n/a	n/a
S981002391		D Beryllium-ICP-Acid Dill.	ug/mL	101.6	<5.00e-03	55.00e-03	55.00e-3	n/a	n/a	n/a	5.00e-03	n/a	n/a
S981002391		D Gadolinium-ICP-Acid Dill.	ug/mL	100.4	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Cobalt-ICP-Acid Dill.	ug/mL	99.80	<2.00e-02	<2.00e-02	<2.00e-2	n/a	n/a	n/a	2.00e-02	n/a	n/a
S981002391		D Copper-ICP-Acid Dill.	ug/mL	94.60	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Lanthanum-ICP-Acid Dill.	ug/mL	100.4	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Magnesium-ICP-Acid Dill.	ug/mL	100.8	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Neodymium-ICP-Acid Dill.	ug/mL	100.0	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Neodymium-ICP-Acid Dill.	ug/mL	99.40	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Phosphorus-ICP-Acid Dill.	ug/mL	99.40	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Lead-ICP-Acid Dill.	ug/mL	98.00	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Sulfur-ICP-Acid Dill.	ug/mL	98.40	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Antimony-ICP-Acid Dill.	ug/mL	101.0	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Selenium-ICP-Acid Dill.	ug/mL	96.40	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Strontium-ICP-Acid Dill.	ug/mL	100.2	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Strontium-ICP-Acid Dill.	ug/mL	98.80	<1.00e-02	<1.00e-02	<1.00e-2	n/a	n/a	n/a	1.00e-02	n/a	n/a
S981002391		D Titanium-ICP-Acid Dill.	ug/mL	99.80	<1.00e-01	<1.00e-01	<1.00e-1	n/a	n/a	n/a	1.00e-01	n/a	n/a
S981002391		D Vanadium-ICP-Acid Dill.	ug/mL	97.60	<2.00e-01	<2.00e-01	<2.00e-1	n/a	n/a	n/a	2.00e-01	n/a	n/a
S981002391		D Zinc-ICP-Acid Dill.	ug/mL	98.80	<5.00e-02	<5.00e-02	<5.00e-2	n/a	n/a	n/a	5.00e-02	n/a	n/a
S981002391		D Acetate by IC-Dioxex 4300/4500	ug/mL	93.50	<2.00e-02	<5.80e-02	<5.80e-2	n/a	n/a	n/a	5.80e-02	n/a	n/a
S981002391		D Glycolate-IC-Dioxex 4300/4500	ug/mL	95.49	<2.00e-02	<5.60e-02	<5.60e-2	n/a	n/a	n/a	5.60e-02	n/a	n/a
S981002391		D Cobalt-60 by BEA	uCi/mL	107.0	<2.48e-05	<2.62e-05	<2.84e-5	n/a	n/a	n/a	2.62e-05	n/a	n/a

5067

25-mar-1999 09:41:07
A-0002-1

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 246
SEGMENT #: N/A

SEGMENT PORTION: Hydrostatic Head Fluid

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002392	D	Silver-TCP-Acid DIL.	ug/mL	98.80	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S98T002392	D	Arsenic-TCP-Acid DIL.	ug/mL	102.2	<1.00e-01	<1.000	<1.00e0	N/A	N/A	N/A	1.000	N/A
S98T002392	D	Beryllium-TCP-Acid DIL.	ug/mL	102.6	<5.00e-03	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002392	D	Cadmium-TCP-Acid DIL.	ug/mL	101.6	<5.00e-03	<5.00e-02	<5.00e-2	N/A	N/A	N/A	5.00e-02	N/A
S98T002392	D	Cerium-TCP-Acid DIL.	ug/mL	100.4	<1.00e-01	<1.000	<1.00e0	N/A	N/A	N/A	1.000	N/A
S98T002392	D	Cobalt-TCP-Acid DIL.	ug/mL	99.80	<2.00e-02	<2.00e-01	<2.00e-1	N/A	N/A	N/A	2.00e-01	N/A
S98T002392	D	Copper-TCP-Acid DIL.	ug/mL	94.60	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S98T002392	D	Lanthanum-TCP-Acid DIL.	ug/mL	100.4	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S98T002392	D	Magnesium-TCP-Acid DIL.	ug/mL	100.8	<1.00e-01	<1.000	<1.00e0	N/A	N/A	N/A	1.000	N/A
S98T002392	D	Molybdenum-TCP-Acid DIL.	ug/mL	100.0	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A
S98T002392	D	Neodymium-TCP-Acid DIL.	ug/mL	99.40	<2.00e-01	<2.000	<2.00e0	N/A	N/A	N/A	2.000	N/A
S98T002392	D	Phosphorus-TCP-Acid DIL.	ug/mL	98.00	<1.00e-01	<1.000	<1.00e0	N/A	N/A	N/A	1.000	N/A
S98T002392	D	Sulfur-TCP-Acid DIL.	ug/mL	98.00	<1.00e-01	<1.000	<1.00e0	1.645	0.61	N/A	1.000	N/A
S98T002392	D	Antimony-TCP-Acid DIL.	ug/mL	101.0	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	6.00e-01	N/A
S98T002392	D	Selenium-TCP-Acid DIL.	ug/mL	96.40	<1.00e-01	<1.000	<1.00e0	N/A	N/A	N/A	1.000	N/A
S98T002392	D	Samarium-TCP-Acid DIL.	ug/mL	100.2	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S98T002392	D	Strontium-TCP-Acid DIL.	ug/mL	98.80	<1.00e-02	<1.00e-01	<1.00e-1	N/A	N/A	N/A	1.00e-01	N/A
S98T002392	D	Titanium-TCP-Acid DIL.	ug/mL	99.80	<2.00e-01	<2.000	<2.00e0	N/A	N/A	N/A	2.000	N/A
S98T002392	D	Thallium-TCP-Acid DIL.	ug/mL	97.60	<2.00e-01	<2.000	<2.00e0	N/A	N/A	N/A	2.000	N/A
S98T002392	D	Vanadium-TCP-Acid DIL.	ug/mL	98.80	<5.00e-02	<5.00e-01	<5.00e-1	N/A	N/A	N/A	5.00e-01	N/A

5068

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 2468
SEGMENT #: 2K

SEGMENT PORTION: Drinkable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T002814	D	Silver-IC-Acid D11.	ug/mL	98.60	<1.00e-02	14.50	46.10	14.35	N/A	6.010	N/A
S98T002814	D	Arsenic-IC-Acid D11.	ug/mL	101.0	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	60.10	N/A
S98T002814	D	Beryllium-IC-Acid D11.	ug/mL	101.5	<5.00e-03	< 3.010	<3.010	N/A	N/A	3.000	N/A
S98T002814	D	Cadmium-IC-Acid D11.	ug/mL	101.5	<5.00e-03	< 3.010	<3.010	N/A	N/A	3.000	N/A
S98T002814	D	Cerium-IC-Acid D11.	ug/mL	100.2	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	60.10	N/A
S98T002814	D	Cobalt-IC-Acid D11.	ug/mL	99.00	<2.00e-02	< 12.00	<1.20e1	N/A	N/A	12.00	N/A
S98T002814	D	Copper-IC-Acid D11.	ug/mL	106.80	<1.00e-02	< 60.10	<6.01e1	N/A	N/A	60.10	N/A
S98T002814	D	Lead-IC-Acid D11.	ug/mL	102.4	<1.00e-02	< 30.10	<3.01e1	N/A	N/A	30.10	N/A
S98T002814	D	Magnesium-IC-Acid D11.	ug/mL	102.4	<1.00e-02	< 30.10	<3.01e1	N/A	N/A	30.10	N/A
S98T002814	D	Molybdenum-IC-Acid D11.	ug/mL	97.40	<1.00e-02	< 60.10	<6.01e1	30.95	0.52	60.10	N/A
S98T002814	D	Nickel-IC-Acid D11.	ug/mL	101.4	<1.00e-01	7.60e+02	339.0	549.5	N/A	120.0	N/A
S98T002814	D	Phosphorus-IC-Acid D11.	ug/mL	100.2	<1.00e-01	43.50	43.70	63.60	N/A	60.10	N/A
S98T002814	D	Silver-IC-Acid D11.	ug/mL	98.20	<1.00e-01	8.00e+02	798.0	799.0	0.25	60.10	N/A
S98T002814	D	Antimony-IC-Acid D11.	ug/mL	101.5	<5.00e-02	< 36.10	<3.61e1	N/A	N/A	36.10	N/A
S98T002814	D	Selenium-IC-Acid D11.	ug/mL	97.00	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	60.10	N/A
S98T002814	D	Samarium-IC-Acid D11.	ug/mL	97.80	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	60.10	N/A
S98T002814	D	Strontium-IC-Acid D11.	ug/mL	97.80	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	6.010	N/A
S98T002814	D	Titanium-IC-Acid D11.	ug/mL	101.4	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	6.010	N/A
S98T002814	D	Vanadium-IC-Acid D11.	ug/mL	96.00	<2.00e-01	<1.20e+02	<1.20e2	N/A	N/A	120.0	N/A
S98T002814	D	Vanadium-IC-Acid D11.	ug/mL	99.40	<5.00e-02	< 30.10	<3.01e1	N/A	N/A	30.10	N/A
S98T002814	D	Acetate by IC-Dionex 4000/4500	ug/mL	100.9	<2.00e-02	3.62e+03	3.62e+03	3.62e+03	0.00	33.62	64.44
S98T002814	D	Glycolate-IC-Dionex 4500 ORGAC	ug/mL	100.9	<2.00e-02	< 62.22	<6.22e1	N/A	N/A	105.6	62.22
S98T002814	D	Cobalt-60 by GEA	uCi/mL	106.4	<1.67e-05	<3.10e-3	<3.10e-3	N/A	N/A	3.00e-03	N/A
S98T002822	D	TIC by Acid/Coulometry	ug/mL	90.03	<3.400	7.68e+03	7.72e+03	7.72e+03	0.91	N/A	5.000
S98T002822	D	Sr 117 by ICP/MS Acid Addition	ug/mL	88.44	<1.50e-04	< 6.075	<6.08e0	N/A	N/A	6.075	N/A
S98T002822	D	Cs 133 by ICP/MS Acid Addition	ug/mL	107.5	<2.00e-06	5.164	5.320	5.242	3.05	N/A	8.10e-02
S98T002822	D	Cs 135 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	1.442	1.490	1.466	3.41	N/A	8.10e-02
S98T002822	D	Cs 137 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	1.150	1.190	1.170	3.42	N/A	8.10e-02
S98T002822	A	Protactinium-231 by ICP/MS	ug/mL	N/A	<1.54e-05	<6.25e-01	<6.25e-1	N/A	N/A	6.25e-01	N/A
S98T002822	A	Thorium-229 by ICP/MS	ug/mL	N/A	N/A	< 1.268	<1.27e0	N/A	N/A	1.268	N/A
S98T002822	A	Thorium-230 by ICP/MS	ug/mL	N/A	N/A	<8.70e-01	<8.69e-1	N/A	N/A	8.70e-01	N/A
S98T002822	A	Thorium-232 by ICP/MS	ug/mL	96.10	N/A	<7.86e-01	<7.86e-1	N/A	N/A	7.86e-01	N/A
S98T002822	A	Pu/Am-241 by ICP/MS	ug/mL	0.25	<1.58e-05	<6.41e-01	<6.41e-1	N/A	N/A	6.41e-01	N/A
S98T002822	A	Pu/Am-243 by ICP/MS	ug/mL	N/A	<1.44e-05	<5.83e-01	<5.83e-1	N/A	N/A	5.83e-01	N/A
S98T002822	A	Pu 244/Am 244 by ICP/MS	ug/mL	N/A	<1.22e-05	<4.96e-01	<4.96e-1	N/A	N/A	4.96e-01	N/A

Appendix A. Opportunistic Analyte Results.
8Y-105 (2)

CORE NUMBER: 246R
SEGMENT #: 3R

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002815	D	Silver--CP-Acid Dil.	ug/mL	98.40	<1.00e-02	13.30	16.00	13.65	5.13	n/a	6.010	n/a
S98T002815	D	Barium--CP-Acid Dil.	ug/mL	101.0	<1.00e-01	< 4.0	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Beryllium--CP-Acid Dil.	ug/mL	101.0	<5.00e-03	< 3.010	<5.010	n/a	n/a	n/a	3.000	n/a
S98T002815	D	Cadmium--CP-Acid Dil.	ug/mL	101.0	<5.00e-03	< 3.010	<5.010	n/a	n/a	n/a	3.000	n/a
S98T002815	D	Cerium--CP-Acid Dil.	ug/mL	100.2	<1.00e-01	< 3.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Cobalt--CP-Acid Dil.	ug/mL	99.00	<2.00e-02	< 12.00	<1.20e1	n/a	n/a	n/a	12.00	n/a
S98T002815	D	Copper--CP-Acid Dil.	ug/mL	94.40	<1.00e-02	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Lead--CP-Acid Dil.	ug/mL	100.6	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	n/a	30.10	n/a
S98T002815	D	Magnesium--CP-Acid Dil.	ug/mL	102.4	<1.00e-01	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Manganese--CP-Acid Dil.	ug/mL	101.0	<5.00e-02	< 31.50	<3.15e1	31.55	0.96	n/a	30.10	n/a
S98T002815	D	Molybdenum--CP-Acid Dil.	ug/mL	102.4	<1.00e-01	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Nickel--CP-Acid Dil.	ug/mL	97.40	<2.00e-01	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Phosphorus--CP-Acid Dil.	ug/mL	101.6	<1.00e-01	4.07e+02	4.09e1	408.0	0.49	n/a	120.0	n/a
S98T002815	D	Selenium--CP-Acid Dil.	ug/mL	100.4	<1.00e-01	80.80	<8.08e1	74.95	15.6	n/a	60.10	n/a
S98T002815	D	Lead--CP-Acid Dil.	ug/mL	98.20	<1.00e-01	8.38e+02	8.45e1	841.5	0.83	n/a	60.10	n/a
S98T002815	D	Antimony--CP-Acid Dil.	ug/mL	101.6	<5.00e-02	< 36.10	<3.61e1	n/a	n/a	n/a	36.10	n/a
S98T002815	D	Selenium--CP-Acid Dil.	ug/mL	97.00	<1.00e-01	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Samarium--CP-Acid Dil.	ug/mL	97.80	<1.00e-02	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Strontium--CP-Acid Dil.	ug/mL	101.4	<1.00e-02	< 6.010	<6.010	n/a	n/a	n/a	6.010	n/a
S98T002815	D	Titanium--CP-Acid Dil.	ug/mL	96.00	<2.00e-01	<1.20e+02	<1.20e2	n/a	n/a	n/a	120.0	n/a
S98T002815	D	Vanadium--CP-Acid Dil.	ug/mL	99.40	<5.00e-02	< 30.10	<3.01e1	n/a	n/a	n/a	30.10	n/a
S98T002815	D	Acetate by IC-Dionex 4000/L500	ug/mL	91.38	3.30e-02	2.25e+03	2.41e+03	2.33e+03	6.87	n/a	35.15	n/a
S98T002815	D	Glycolate--IC-Dionex 4500 OHGAC	ug/mL	88.80	2.30e-02	<3.53e-03	<3.53e-3	n/a	n/a	n/a	3.00e-03	n/a
S98T002815	D	Cobalt-60 by GEA	uCi/mL	106.4	<1.67e-05	< 33.54	<3.39e1	n/a	n/a	n/a	33.54	n/a
S98T002823	A	TIC by Acid/Countmetry	ug/mL	90.03	3.400	5.67e+03	5.53e+03	5.60e+03	2.50	n/a	5.000	n/a
S98T002823	D	Sr 117 by ICP/MS Acid Addition	ug/mL	88.44	<1.50e-04	< 6.075	<6.080	n/a	n/a	n/a	6.075	n/a
S98T002823	D	Cs 133 by ICP/MS Acid Addition	ug/mL	107.5	<2.00e-06	1.624	1.630	1.627	0.62	n/a	8.10e-02	n/a
S98T002823	D	Cs 135 by ICP/MS Acid Addition	ug/mL	n/a	<2.00e-06	1.397	1.390	1.393	0.72	n/a	8.10e-02	n/a
S98T002823	A	Protactinium-231 by ICP/MS	ug/mL	n/a	<1.54e-05	<6.25e-01	<6.25e-1	n/a	n/a	n/a	6.25e-01	n/a
S98T002823	A	Thorium-229 by ICP/MS	ug/mL	n/a	n/a	< 1.268	<1.270	n/a	n/a	n/a	1.268	n/a
S98T002823	A	Thorium-230 by ICP/MS	ug/mL	n/a	n/a	<8.70e-01	<8.69e-1	n/a	n/a	n/a	8.70e-01	n/a
S98T002823	A	Thorium-232 by ICP/MS	ug/mL	96.10	n/a	<7.86e-01	<7.86e-1	n/a	n/a	n/a	7.86e-01	n/a
S98T002823	A	Pu 241 by ICP/MS	ug/mL	0.23	<1.59e-05	<6.41e-01	<6.41e-1	n/a	n/a	n/a	6.41e-01	n/a
S98T002823	A	Pu 243 by ICP/MS	ug/mL	n/a	<1.44e-05	<5.83e-01	<5.83e-1	n/a	n/a	n/a	5.83e-01	n/a
S98T002823	A	Pu 244 by ICP/MS	ug/mL	n/a	<1.22e-05	<4.96e-01	<4.96e-1	n/a	n/a	n/a	4.96e-01	n/a

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 246R
SEGMENT #: 4R

SEGMENT PORTION: Drainable Liquid

Sample #	B #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err %
S98T002316	D	Silver-IC-Acid Dil.	ug/mL	99.60	< 1.00e-03	6.30	12.81	13.55	11.1	N/A	6.010	N/A
S98T002316	D	Arsenic-IC-Acid Dil.	ug/mL	100.0	< 1.00e-01	40.10	45.01	N/A	N/A	N/A	9.100	N/A
S98T002316	D	Beryllium-IC-Acid Dil.	ug/mL	100.0	< 5.00e-03	3.010	5.0160	N/A	N/A	N/A	3.000	N/A
S98T002316	D	Cadmium-IC-Acid Dil.	ug/mL	102.4	< 3.00e-01	40.10	45.01	N/A	N/A	N/A	3.000	N/A
S98T002316	D	Cerium-IC-Acid Dil.	ug/mL	99.80	< 1.00e-01	40.10	45.01	N/A	N/A	N/A	12.000	N/A
S98T002316	D	Cobalt-IC-Acid Dil.	ug/mL	99.20	< 2.00e-02	12.00	41.2041	N/A	N/A	N/A	6.010	N/A
S98T002316	D	Copper-IC-Acid Dil.	ug/mL	92.00	< 2.00e-02	40.10	45.01	N/A	N/A	N/A	30.10	N/A
S98T002316	D	Lead-IC-Acid Dil.	ug/mL	100.2	< 5.00e-03	30.10	45.01	N/A	N/A	N/A	40.10	N/A
S98T002316	D	Mercury-IC-Acid Dil.	ug/mL	100.4	< 1.00e-01	40.10	45.01	N/A	N/A	N/A	30.10	N/A
S98T002316	D	Nickel-IC-Acid Dil.	ug/mL	100.4	< 5.00e-02	35.40	45.01	35.15	1.72	N/A	40.10	N/A
S98T002316	D	Phosphorus-IC-Acid Dil.	ug/mL	97.60	< 1.00e-01	40.10	45.01	418.0	2.94	N/A	120.0	N/A
S98T002316	D	Selenium-IC-Acid Dil.	ug/mL	100.2	< 2.00e-01	4.00e+02	418.0	412.0	2.94	N/A	40.10	N/A
S98T002316	D	Silver-IC-Acid Dil.	ug/mL	100.4	< 1.00e-01	75.36	75.36	75.25	0.13	N/A	60.10	N/A
S98T002316	D	Sulfur-IC-Acid Dil.	ug/mL	98.60	< 1.00e-02	8.95e+02	895.0	895.0	0.00	N/A	36.10	N/A
S98T002316	D	Titanium-IC-Acid Dil.	ug/mL	101.8	< 6.00e-02	36.10	45.01	N/A	N/A	N/A	60.10	N/A
S98T002316	D	Vanadium-IC-Acid Dil.	ug/mL	96.60	< 1.00e-01	40.10	45.01	N/A	N/A	N/A	60.10	N/A
S98T002316	D	Strontium-IC-Acid Dil.	ug/mL	97.80	< 1.00e-02	6.010	45.0160	N/A	N/A	N/A	6.010	N/A
S98T002316	D	Titanium-IC-Acid Dil.	ug/mL	101.2	< 1.00e-02	6.010	45.0160	N/A	N/A	N/A	6.010	N/A
S98T002316	D	Thallium-IC-Acid Dil.	ug/mL	96.40	< 2.00e-01	41.20e+02	41.2042	N/A	N/A	N/A	120.0	N/A
S98T002316	D	Vanadium-IC-Acid Dil.	ug/mL	99.40	< 5.00e-02	30.10	45.01	1.34e+03	4.48	N/A	298.8	N/A
S98T002316	D	Acetate by IC-Dionex 4000/4500	ug/mL	92.52	< 2.00e-02	1.38e+03	1.31e+03	1.34e+03	4.48	N/A	298.8	N/A
S98T002316	D	Glycolate by IC-Dionex 4500 ORGAC	ug/mL	95.05	< 2.00e-02	2.88e+02	2.88e+02	2.88e+02	N/A	N/A	3.00e-03	N/A
S98T002316	D	Cobalt-60 by GEA	uCi/mL	106.4	< 1.67e-05	5.29e+03	5.23e+03	5.26e+03	1.14	N/A	5.000	N/A
S98T002824	A	TIC by Acid/Coulometry	ug/mL	99.67	< 7.00e-01	6.075	46.0840	N/A	N/A	N/A	6.075	N/A
S98T002824	D	Sn 117 by ICP/MS Acid Addition	ug/mL	81.82	< 1.50e-04	7.270	7.270	7.213	1.52	N/A	8.10e-02	N/A
S98T002824	D	Cs 133 by ICP/MS Acid Addition	ug/mL	105.5	< 1.24e-05	2.045	2.060	2.053	0.98	N/A	8.10e-02	N/A
S98T002824	D	Cs 135 by ICP/MS Acid Addition	ug/mL	N/A	< 2.00e-06	1.867	1.910	1.889	2.12	N/A	8.10e-02	N/A
S98T002824	D	Cs 137 by ICP/MS Acid Addition	ug/mL	N/A	< 2.00e-06	1.867	1.910	1.889	2.12	N/A	8.10e-02	N/A
S98T002824	A	Protactinium-231 by ICP/MS	ug/mL	N/A	< 1.54e-05	46.25e-01	46.25e-01	N/A	N/A	N/A	6.25e-01	N/A
S98T002824	A	Thorium-229 by ICP/MS	ug/mL	N/A	N/A	< 1.268	41.2760	N/A	N/A	N/A	1.268	N/A
S98T002824	A	Thorium-230 by ICP/MS	ug/mL	N/A	N/A	< 8.70e-01	48.69e-1	N/A	N/A	N/A	8.70e-01	N/A
S98T002824	A	Thorium-232 by ICP/MS	ug/mL	96.10	< 7.86e-01	47.86e-1	47.86e-1	N/A	N/A	N/A	7.86e-01	N/A
S98T002824	A	Pu/Am-241 by ICP/MS	ug/mL	0.25	< 1.58e-05	46.41e-01	46.41e-01	N/A	N/A	N/A	6.41e-01	N/A
S98T002824	A	Am-243/Am-244 by ICP/MS	ug/mL	N/A	< 1.44e-05	45.83e-01	45.83e-01	N/A	N/A	N/A	5.83e-01	N/A
S98T002824	A	Pu-244/Am-244 by ICP/MS	ug/mL	N/A	< 1.22e-05	44.96e-01	44.96e-01	N/A	N/A	N/A	4.96e-01	N/A

Appendix A. Opportunistic Analyte Results.
8Y-105 (2)CORE NUMBER: 24GR
SEGMENT #: 5R

SEGMENT PORTION: Drainable Liquid

Sample#	P	# Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPP % Spk Rec %	Det Limit	Count Error
S98T002817	D	Silver-CP-Acid D11	ug/mL	99.60	< 1.00e-02	13.30	13.80	13.55	3.69	6.010	D/A
S98T002817	D	Antimony-CP-Acid D11	ug/mL	101.6	< 1.00e-01	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Barium-CP-Acid D11	ug/mL	100.6	< 1.00e-03	< 3.010	< 3.010	D/A	D/A	3.000	D/A
S98T002817	D	Beryllium-CP-Acid D11	ug/mL	101.6	< 1.00e-03	< 3.010	< 3.010	D/A	D/A	3.000	D/A
S98T002817	D	Cadmium-CP-Acid D11	ug/mL	99.80	< 1.00e-01	< 3.010	< 3.010	D/A	D/A	60.10	D/A
S98T002817	D	Cesium-CP-Acid D11	ug/mL	99.80	< 1.00e-01	< 12.00	< 12.00	D/A	D/A	12.00	D/A
S98T002817	D	Cobalt-CP-Acid D11	ug/mL	99.20	< 1.00e-02	< 6.010	< 6.010	D/A	D/A	6.010	D/A
S98T002817	D	Copper-CP-Acid D11	ug/mL	94.00	< 1.00e-02	< 6.010	< 6.010	D/A	D/A	30.10	D/A
S98T002817	D	Lead-CP-Acid D11	ug/mL	100.6	< 1.00e-01	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Mercury-CP-Acid D11	ug/mL	100.6	< 1.00e-02	< 36.60	< 36.60	D/A	D/A	30.10	D/A
S98T002817	D	Molybdenum-CP-Acid D11	ug/mL	101.6	< 1.00e-01	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Nickel-CP-Acid D11	ug/mL	97.60	< 1.00e-02	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Phosphorus-CP-Acid D11	ug/mL	101.6	< 1.00e-01	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Selenium-CP-Acid D11	ug/mL	101.6	< 1.00e-01	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Silver-CP-Acid D11	ug/mL	99.60	< 1.00e-02	< 36.10	< 36.10	D/A	D/A	36.10	D/A
S98T002817	D	Antimony-CP-Acid D11	ug/mL	101.6	< 1.00e-02	< 40.10	< 40.10	D/A	D/A	60.10	D/A
S98T002817	D	Barium-CP-Acid D11	ug/mL	97.60	< 1.00e-01	< 60.10	< 60.10	D/A	D/A	60.10	D/A
S98T002817	D	Beryllium-CP-Acid D11	ug/mL	97.80	< 1.00e-02	< 6.010	< 6.010	D/A	D/A	6.010	D/A
S98T002817	D	Cadmium-CP-Acid D11	ug/mL	101.2	< 1.00e-02	< 20.00	< 20.00	D/A	D/A	60.10	D/A
S98T002817	D	Cesium-CP-Acid D11	ug/mL	96.40	< 1.00e-02	< 30.10	< 30.10	D/A	D/A	30.10	D/A
S98T002817	D	Cobalt-CP-Acid D11	ug/mL	99.40	< 1.00e-02	< 1.56e+03	< 1.56e+03	D/A	D/A	298.8	D/A
S98T002817	D	Copper-CP-Acid D11	ug/mL	95.05	< 1.00e-02	< 2.88e+02	< 2.88e+02	D/A	D/A	288.5	D/A
S98T002817	D	Lead-CP-Acid D11	ug/mL	106.4	< 1.00e-05	< 3.57e-03	< 3.57e-03	D/A	D/A	4.00e-03	D/A
S98T002817	D	Mercury-CP-Acid D11	ug/mL	99.67	< 1.00e-02	< 6.075	< 6.075	D/A	D/A	5.000	D/A
S98T002817	D	Nickel-CP-Acid D11	ug/mL	81.82	< 1.00e-04	< 6.229	< 6.229	D/A	D/A	6.075	D/A
S98T002817	D	Phosphorus-CP-Acid D11	ug/mL	105.5	< 1.00e-06	< 1.827	< 1.827	D/A	D/A	8.10e-02	D/A
S98T002817	D	Selenium-CP-Acid D11	ug/mL	99.40	< 1.00e-06	< 1.628	< 1.628	D/A	D/A	8.10e-02	D/A
S98T002817	D	Silver-CP-Acid D11	ug/mL	99.67	< 1.00e-05	< 4.25e-01	< 4.25e-01	D/A	D/A	6.25e-01	D/A
S98T002817	D	Antimony-CP-Acid D11	ug/mL	99.67	< 1.00e-05	< 1.268	< 1.268	D/A	D/A	1.268	D/A
S98T002817	D	Barium-CP-Acid D11	ug/mL	99.67	< 1.00e-01	< 48.69e-1	< 48.69e-1	D/A	D/A	8.70e-01	D/A
S98T002817	D	Beryllium-CP-Acid D11	ug/mL	96.10	< 1.00e-05	< 7.86e-1	< 7.86e-1	D/A	D/A	7.86e-01	D/A
S98T002817	D	Cadmium-CP-Acid D11	ug/mL	0.25	< 1.00e-05	< 4.61e-01	< 4.61e-01	D/A	D/A	6.41e-01	D/A
S98T002817	D	Cesium-CP-Acid D11	ug/mL	99.67	< 1.00e-05	< 5.83e-01	< 5.83e-01	D/A	D/A	5.83e-01	D/A
S98T002817	D	Cobalt-CP-Acid D11	ug/mL	99.67	< 1.00e-05	< 4.96e-01	< 4.96e-01	D/A	D/A	4.96e-01	D/A

5072

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 246R
SEGMENT #: 48

SEGMENT PORTION: Draineable Liquid

Sample #	P.A.#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	Rpb %	Spk Rec %	Det Limit	Count	Err%
S98T002818	D	Silver-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-02	< 15.50	15.50	15.05	1.92	n/a	6.010	n/a	n/a
S98T002818	D	Arsenic-ICP-Acid Dil.	ug/mL	100.4	< 1.00e-01	< 40.10	< 40.10	n/a	n/a	n/a	3.000	n/a	n/a
S98T002818	D	Beryllium-ICP-Acid Dil.	ug/mL	102.0	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	n/a	3.000	n/a	n/a
S98T002818	D	Cadmium-ICP-Acid Dil.	ug/mL	102.2	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	n/a	3.000	n/a	n/a
S98T002818	D	Cerium-ICP-Acid Dil.	ug/mL	99.80	< 1.00e-01	< 40.10	< 40.10	n/a	n/a	n/a	12.00	n/a	n/a
S98T002818	D	Cobalt-ICP-Acid Dil.	ug/mL	99.20	< 2.00e-02	< 12.00	< 12.00	n/a	n/a	n/a	6.010	n/a	n/a
S98T002818	D	Copper-ICP-Acid Dil.	ug/mL	94.00	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	n/a	30.10	n/a	n/a
S98T002818	D	Lead-ICP-Acid Dil.	ug/mL	100.4	< 5.00e-03	< 30.10	< 30.10	n/a	n/a	n/a	6.010	n/a	n/a
S98T002818	D	Magnesium-ICP-Acid Dil.	ug/mL	103.8	< 1.00e-01	< 40.10	< 40.10	n/a	n/a	n/a	30.10	n/a	n/a
S98T002818	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.6	< 5.00e-02	< 35.60	< 35.60	35.95	1.95	n/a	6.010	n/a	n/a
S98T002818	D	Nickel-ICP-Acid Dil.	ug/mL	97.80	< 1.00e-01	< 40.10	< 40.10	n/a	n/a	n/a	6.010	n/a	n/a
S98T002818	D	Phosphorus-ICP-Acid Dil.	ug/mL	102.4	< 2.00e-01	< 5.90e+02	< 5.90e+02	457.0	0.42	n/a	120.0	n/a	n/a
S98T002818	D	Lead-ICP-Acid Dil.	ug/mL	101.0	< 1.00e-01	< 67.90	< 67.90	68.50	1.75	n/a	60.10	n/a	n/a
S98T002818	D	Sulfur-ICP-Acid Dil.	ug/mL	97.40	< 1.00e-01	< 24.00	< 24.00	24.00	0.65	n/a	60.10	n/a	n/a
S98T002818	D	Antimony-ICP-Acid Dil.	ug/mL	101.8	< 5.00e-02	< 36.10	< 36.10	n/a	n/a	n/a	36.10	n/a	n/a
S98T002818	D	Selenium-ICP-Acid Dil.	ug/mL	97.80	< 1.00e-01	< 40.10	< 40.10	n/a	n/a	n/a	60.10	n/a	n/a
S98T002818	D	Strontium-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-01	< 60.10	< 60.10	n/a	n/a	n/a	60.10	n/a	n/a
S98T002818	D	Titanium-ICP-Acid Dil.	ug/mL	101.0	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	n/a	6.010	n/a	n/a
S98T002818	D	Vanadium-ICP-Acid Dil.	ug/mL	95.60	< 5.00e-02	< 30.10	< 30.10	n/a	n/a	n/a	120.0	n/a	n/a
S98T002818	D	Vanadium-ICP-Acid Dil.	ug/mL	99.40	< 5.00e-02	< 1.52e+03	< 1.52e+03	1.53e+03	0.00	n/a	123.0	n/a	n/a
S98T002818	D	Acetate by IC-Dionex 4000/4500	ug/mL	92.52	< 2.00e-02	< 1.19e+02	< 1.19e+02	n/a	n/a	n/a	118.8	n/a	n/a
S98T002818	D	Glycolate by IC-Dionex 4500 ORGAC	ug/mL	95.05	< 2.00e-02	< 4.00e+03	< 4.00e+03	3.77e+3	0.00	n/a	4.00e-03	n/a	n/a
S98T002818	D	Cobalt-60 by GEA	uCi/mL	103.3	< 1.24e-03	< 3.72e+03	< 3.72e+03	3.88e+03	4.21	n/a	5.000	n/a	n/a
S98T002826	D	TIC by Acid/Coulometry	ug/mL	99.67	< 7.00e-01	< 6.075	< 6.075	< 6.080	n/a	n/a	8.10e-02	n/a	n/a
S98T002826	D	Sn 117 by ICP/MS Acid Addition	ug/mL	105.5	< 1.50e-04	< 2.122	< 2.122	2.110	0.17	n/a	8.10e-02	n/a	n/a
S98T002826	D	Cs 133 by ICP/MS Acid Addition	ug/mL	89.00	< 2.00e-06	< 2.074	< 2.074	2.092	1.91	n/a	8.10e-02	n/a	n/a
S98T002826	D	Cs 135 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	< 2.074	< 2.074	2.110	1.91	n/a	8.10e-02	n/a	n/a
S98T002826	D	Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	< 2.074	< 2.074	2.110	1.91	n/a	8.10e-02	n/a	n/a
S98T002826	A	Protactinium-231 by ICP/MS	ug/mL	n/a	< 5.57e-06	< 2.26e-01	< 2.26e-01	n/a	n/a	n/a	2.26e-01	n/a	n/a
S98T002826	A	Thorium-229 by ICP/MS	ug/mL	n/a	< 4.46e-06	< 1.81e-01	< 1.81e-01	n/a	n/a	n/a	1.81e-01	n/a	n/a
S98T002826	A	Thorium-230 by ICP/MS	ug/mL	n/a	< 5.5e-06	< 1.44e-01	< 1.44e-01	n/a	n/a	n/a	1.44e-01	n/a	n/a
S98T002826	A	Thorium-232 by ICP/MS	ug/mL	99.20	< 1.93e-06	< 4.82e-02	< 4.82e-02	n/a	n/a	n/a	7.80e-02	n/a	n/a
S98T002826	A	Pu/Am-241 by ICP/MS	ug/mL	109.7	< 3.71e-06	< 1.50e-01	< 1.50e-01	n/a	n/a	n/a	1.50e-01	n/a	n/a
S98T002826	A	Am-243 by ICP/MS	ug/mL	109.7	< 3.83e-06	< 1.52e-01	< 1.52e-01	n/a	n/a	n/a	1.52e-01	n/a	n/a
S98T002826	A	Pu-244/Cm-244 by ICP/MS	ug/mL	n/a	< 3.41e-06	< 1.38e-01	< 1.38e-01	n/a	n/a	n/a	1.38e-01	n/a	n/a

Appendix A. Opportunistic Analyte Results.
8Y-105 (2)

CORE NUMBER: 246R
SEGMENT #: 7R

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	%Spk	Rec. %	Det Limit	Count Err%
S98T002819	D	Stilbene-ICP-Acid D11	ug/mL	98.40	< 1.00e-02	15.50	15.70	15.60	1.28	92.70	92.70	6.010	0/8
S98T002819	D	Arsenic-ICP-Acid D11	ug/mL	101.0	< 1.00e-01	< 3.010	< 3.010	0/8	0/8	100.0	100.0	3.000	0/8
S98T002819	D	Beryllium-ICP-Acid D11	ug/mL	101.6	< 5.00e-03	< 3.010	< 3.010	0/8	0/8	100.0	100.0	3.000	0/8
S98T002819	D	Cadmium-ICP-Acid D11	ug/mL	101.6	< 5.00e-03	< 3.010	< 3.010	0/8	0/8	100.0	100.0	3.000	0/8
S98T002819	D	Cerium-ICP-Acid D11	ug/mL	100.2	< 1.00e-01	< 3.010	< 3.010	0/8	0/8	98.70	98.70	40.10	0/8
S98T002819	D	Cobalt-ICP-Acid D11	ug/mL	99.00	< 2.00e-02	< 12.00	< 20.61	0/8	0/8	98.20	98.20	12.00	0/8
S98T002819	D	Copper-ICP-Acid D11	ug/mL	94.40	< 1.00e-02	< 6.010	< 4.010	0/8	0/8	98.70	98.70	5.010	0/8
S98T002819	D	Lead-ICP-Acid D11	ug/mL	100.6	< 5.00e-02	< 6.010	< 3.010	0/8	0/8	98.70	98.70	30.10	0/8
S98T002819	D	Magnesium-ICP-Acid D11	ug/mL	102.4	< 1.00e-01	< 6.010	< 4.010	0/8	0/8	98.70	98.70	40.10	0/8
S98T002819	D	Manganese-ICP-Acid D11	ug/mL	101.0	< 5.00e-02	< 35.30	< 35.40	35.35	0.28	92.80	92.80	30.10	0/8
S98T002819	D	Nickel-ICP-Acid D11	ug/mL	97.60	< 1.00e-01	< 6.010	< 4.010	0/8	0/8	98.40	98.40	40.10	0/8
S98T002819	D	Phosphorus-ICP-Acid D11	ug/mL	101.6	< 2.00e-01	5.09e+02	497.0	503.0	2.39	98.50	98.50	120.0	0/8
S98T002819	D	Lead-ICP-Acid D11	ug/mL	100.4	< 1.00e-01	77.60	80.80	79.20	4.04	99.50	99.50	40.10	0/8
S98T002819	D	Sulfur-ICP-Acid D11	ug/mL	98.20	< 1.00e-01	8.88e+02	883.0	885.5	0.55	95.70	95.70	60.10	0/8
S98T002819	D	Antimony-ICP-Acid D11	ug/mL	101.6	< 5.00e-02	< 36.10	< 3.610	0/8	0/8	101.0	101.0	36.10	0/8
S98T002819	D	Selenium-ICP-Acid D11	ug/mL	97.00	< 1.00e-01	< 6.010	< 6.010	0/8	0/8	96.20	96.20	60.10	0/8
S98T002819	D	Samarium-ICP-Acid D11	ug/mL	97.80	< 1.00e-02	< 6.010	< 6.010	0/8	0/8	95.70	95.70	6.010	0/8
S98T002819	D	Strontium-ICP-Acid D11	ug/mL	101.4	< 1.00e-02	< 6.010	< 6.010	0/8	0/8	99.60	99.60	6.010	0/8
S98T002819	D	Thallium-ICP-Acid D11	ug/mL	99.40	< 5.00e-02	< 1.20e+02	< 1.20e2	0/8	0/8	98.00	98.00	120.0	0/8
S98T002819	D	Vanadium-ICP-Acid D11	ug/mL	95.70	< 2.00e-02	< 30.10	< 3.010	0/8	0/8	98.00	98.00	30.10	0/8
S98T002819	D	Acetate by IC-Dionex 4000/4500	ug/mL	106.9	< 2.00e-02	2.20e+03	2.35e+03	2.28e+03	6.59	98.03	98.03	35.15	0/8
S98T002819	D	Glycolate by IC-Dionex 4500 ORGAC	ug/mL	103.3	< 2.00e-02	< 33.94	< 3.94e1	0/8	0/8	129.7	129.7	33.94	0/8
S98T002827	A	Cobalt-60 by GEA	ug/mL	100.5	< 1.24e-03	< 3.85e-03	< 3.87e-3	3.28e+03	0.30	103.0	103.0	5.000	0/8
S98T002827	A	Sn 117 by ICP/Oxymetry	ug/mL	88.44	< 1.50e-04	< 6.075	< 6.080	0/8	0/8	0/8	0/8	4.00e-03	0/8
S98T002827	D	Cs 133 by ICP/MS Acid Addition	ug/mL	107.5	< 2.00e-06	10.23	10.40	10.32	1.94	106.0	106.0	8.10e-02	0/8
S98T002827	D	Cs 135 by ICP/MS Acid Addition	ug/mL	107.5	< 2.00e-06	3.730	3.660	3.695	1.89	0/8	0/8	8.10e-02	0/8
S98T002827	D	Cs 137 by ICP/MS Acid Addition	ug/mL	0/8	< 2.00e-06	2.730	2.830	2.780	3.60	0/8	0/8	8.10e-02	0/8
S98T002827	A	Protactinium-231 by ICP/MS	ug/mL	0/8	< 1.54e-05	< 6.25e-01	< 6.25e-1	0/8	0/8	0/8	0/8	6.25e-01	0/8
S98T002827	A	Thorium-229 by ICP/MS	ug/mL	0/8	0/8	< 1.268	< 1.270	0/8	0/8	0/8	0/8	1.268	0/8
S98T002827	A	Thorium-230 by ICP/MS	ug/mL	0/8	0/8	< 8.70e-01	< 8.69e-1	0/8	0/8	0/8	0/8	8.70e-01	0/8
S98T002827	A	Thorium-232 by ICP/MS	ug/mL	96.10	0/8	< 7.86e-01	< 7.86e-1	0/8	0/8	81.43	81.43	7.86e-01	0/8
S98T002827	A	Pu Am-241 by ICP/MS	ug/mL	0.25	< 1.58e-05	< 6.41e-01	< 6.41e-1	0/8	0/8	2.42e-01	2.42e-01	6.41e-01	0/8
S98T002827	A	Pu Am-243 by ICP/MS	ug/mL	0/8	< 1.44e-05	< 5.83e-01	< 5.83e-1	0/8	0/8	0/8	0/8	5.83e-01	0/8
S98T002827	A	Pu 244/Am 244 by ICP/MS	ug/mL	0/8	< 1.22e-05	< 4.96e-01	< 4.96e-1	0/8	0/8	0/8	0/8	4.96e-01	0/8

Appendix A. Opportunistic Analyte Results.
BI-105 (2)

CORE NUMBER: 246R
SEGMENT #: 8R

SEGMENT PORTION: L Lower Half of Segment

Sample#	RAL	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk	Rec %	Det Limit	Count	Err%
S981002844	F	Silver - ICP-Fusion	ug/g	100.4	<1.00e-02	<2.07e-02	<2.06e2	n/a	n/a	91.40	207.0	n/a	n/a	n/a
S981002844	F	Arsenic - ICP-Fusion	ug/g	102.4	<1.00e-02	<2.07e-03	<2.06e3	n/a	n/a	101.0	2.07e+03	n/a	n/a	n/a
S981002844	F	Beryllium - ICP-Fusion	ug/g	103.4	<1.00e-03	<1.03e+02	<1.03e2	n/a	n/a	100.0	103.0	n/a	n/a	n/a
S981002844	F	Cadmium - ICP-Fusion	ug/g	103.4	<1.00e-03	<1.03e+02	<1.03e2	n/a	n/a	100.0	103.0	n/a	n/a	n/a
S981002844	F	Cerium - ICP-Fusion	ug/g	102.4	<1.00e-03	<2.07e+03	<2.06e3	n/a	n/a	102.0	2.07e+03	n/a	n/a	n/a
S981002844	F	Cobalt - ICP-Fusion	ug/g	100.4	<1.00e-02	<4.11e2	<4.11e2	n/a	n/a	98.40	414.0	n/a	n/a	n/a
S981002844	F	Copper - ICP-Fusion	ug/g	102.4	<1.00e-02	<2.07e+02	<2.06e2	n/a	n/a	99.40	207.0	n/a	n/a	n/a
S981002844	F	Lanthanum - ICP-Fusion	ug/g	102.4	<1.00e-02	<1.03e+03	<1.03e3	n/a	n/a	100.0	1.03e+03	n/a	n/a	n/a
S981002844	F	Magnesium - ICP-Fusion	ug/g	105.0	<1.00e-02	<2.07e+03	<2.06e3	n/a	n/a	103.0	2.07e+03	n/a	n/a	n/a
S981002844	F	Molybdenum - ICP-Fusion	ug/g	102.4	<1.00e-02	<1.03e+03	<1.03e3	n/a	n/a	101.0	1.03e+03	n/a	n/a	n/a
S981002844	F	Neodymium - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	97.40	2.07e+03	n/a	n/a	n/a
S981002844	F	Phosphorus - ICP-Fusion	ug/g	103.0	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	99.10	2.07e+03	n/a	n/a	n/a
S981002844	F	Lead - ICP-Fusion	ug/g	102.2	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	116.0	1.44e+03	n/a	n/a	n/a
S981002844	F	Sulfur - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	99.70	2.07e+03	n/a	n/a	n/a
S981002844	F	Antimony - ICP-Fusion	ug/g	102.8	<1.00e-02	<1.03e+03	<1.03e3	n/a	n/a	103.0	1.03e+03	n/a	n/a	n/a
S981002844	F	Selenium - ICP-Fusion	ug/g	98.20	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	103.0	2.07e+03	n/a	n/a	n/a
S981002844	F	Strontium - ICP-Fusion	ug/g	100.0	<1.00e-01	<2.07e+03	<2.06e3	n/a	n/a	99.10	2.07e+03	n/a	n/a	n/a
S981002844	F	Tantalum - ICP-Fusion	ug/g	103.2	<1.00e-02	<1.03e+03	<1.03e3	n/a	n/a	101.0	1.03e+03	n/a	n/a	n/a
S981002844	F	Thallium - ICP-Fusion	ug/g	99.00	<1.00e-02	<2.07e+03	<2.06e3	n/a	n/a	97.40	2.07e+03	n/a	n/a	n/a
S981002844	F	Unidentified	ug/g	101.0	<1.00e-02	<2.07e+03	<2.06e3	n/a	n/a	97.40	2.07e+03	n/a	n/a	n/a
S981002844	F	Cobalt-60 By GEA	uCi/g	104.3	<4.14e-02	<2.07e+02	<2.06e2	1.04e+03	2.90	97.40	207.0	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	115.3	8.00e-03	6.14e-2	6.14e-2	n/a	n/a	95.90	4.14e+03	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	99.00	<2.00e-05	2.303	2.303	n/a	n/a	98.80	1.03e+03	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	6.50e-05	4.595	4.595	n/a	n/a	n/a	2.30e-02	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	1.12e-04	4.595	4.595	n/a	n/a	n/a	8.20e-02	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	<4.35e-05	<1.62e-01	<1.62e-1	n/a	n/a	n/a	1.62e-01	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	<4.00e-05	<2.24e-01	<2.24e-1	n/a	n/a	n/a	2.24e-01	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	<1.84e-05	<6.88e-02	<6.88e-2	n/a	n/a	n/a	6.88e-02	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	107.0	<4.12e-05	2.539	2.539	n/a	n/a	99.50	1.54e-01	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	103.9	<1.85e-05	<6.88e-02	<6.88e-2	2.625	7.24	101.0	6.88e-02	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	<8.80e-06	<3.46e-2	<3.46e-2	n/a	n/a	n/a	3.46e-02	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	n/a	<2.35e-05	<2.35e-05	<2.35e-05	n/a	n/a	n/a	2.35e-05	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	86.30	<1.00e-02	16.10	16.10	15.90	2.52	96.20	10.20	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	93.60	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	104.6	102.0	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	102.2	<1.00e-03	5.100	5.100	n/a	n/a	104.0	5.100	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	93.20	<5.00e-03	5.100	5.100	n/a	n/a	104.2	5.100	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	97.20	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	100.8	102.0	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	94.00	<2.00e-02	10.20	10.20	n/a	n/a	102.2	20.40	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	86.60	<1.00e-02	10.20	10.20	n/a	n/a	92.40	10.20	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	97.00	<5.00e-02	1.24e+02	1.24e+02	n/a	n/a	101.2	51.00	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	99.00	<1.00e-01	5.100	5.100	133.0	13.5	104.4	102.0	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	95.60	<5.00e-02	5.100	5.100	n/a	n/a	103.6	51.00	n/a	n/a	n/a
S981002844	F	LA-505159	ug/g	96.00	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	100.4	102.0	n/a	n/a	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002846	A	Phosphorus - ICP-Acid Digest	ug/g	92.60	<2.00e-01	4.44e+03	6.99e+03	5.71e+03	44.5	-1.100e1	204.0	n/a
S98T002846	A	Lead - ICP-Acid Digest	ug/g	90.60	<1.00e-01	<1.02e+02	<1.01e2	<1.01e2	n/a	n/a	110.4	n/a
S98T002846	A	Sulfur - ICP-Acid Digest	ug/g	89.40	<1.50e-01	1.12e+03	1.14e+03	1.14e+03	4.41	96.00	102.0	n/a
S98T002846	A	Antimony - ICP-Acid Digest	ug/g	91.80	<6.00e-02	<6.10	<6.00e1	n/a	n/a	104.0	61.00	n/a
S98T002846	A	Selenium - ICP-Acid Digest	ug/g	92.60	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	104.8	102.0	n/a
S98T002846	A	Samarium - ICP-Acid Digest	ug/g	95.80	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	98.00	102.0	n/a
S98T002846	A	Strontium - ICP-Acid Digest	ug/g	96.00	<1.00e-01	<1.02e+02	<1.01e2	n/a	n/a	95.00	102.0	n/a
S98T002846	A	Titanium - ICP-Acid Digest	ug/g	96.60	<1.00e-02	1.18e+01	1.20e1	1.11e+03	18.50	101.2	10.20	n/a
S98T002846	A	Thallium - ICP-Acid Digest	ug/g	89.00	<2.00e-01	<2.04e+02	<2.03e2	n/a	n/a	97.00	204.0	n/a
S98T002846	A	Vanadium - ICP-Acid Digest	ug/g	93.60	<5.00e-02	<5.10	<5.05e1	n/a	n/a	100.2	51.00	n/a
S98T002846	W	Acetate by IC-DIOMEX 4000/4500	ug/g	93.10	3.20e-02	6.61e+02	6.77e+02	669.0	n/a	n/a	82.22	n/a
S98T002846	W	Glycolate IC-DIOMEX 4000/4500	ug/g	91.76	<5.60e-02	<2.30e+02	<2.30e2	n/a	n/a	94.44	230.2	n/a

Drainable Liquids: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002820	D	Silver-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	16.20	15.90	16.05	1.87	n/a	6.010	n/a
S98T002820	D	Arsenic-ICP-Acid Dil.	ug/mL	100.4	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T002820	D	Beryllium-ICP-Acid Dil.	ug/mL	102.0	<5.00e-03	<3.010	<3.01e0	n/a	n/a	n/a	3.000	n/a
S98T002820	D	Calcium-ICP-Acid Dil.	ug/mL	102.2	<5.00e-03	<3.010	<3.01e0	n/a	n/a	n/a	3.000	n/a
S98T002820	D	Cerium-ICP-Acid Dil.	ug/mL	99.80	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T002820	D	Cobalt-ICP-Acid Dil.	ug/mL	99.20	<2.00e-02	<12.00	<12.0e1	n/a	n/a	n/a	12.00	n/a
S98T002820	D	Copper-ICP-Acid Dil.	ug/mL	94.00	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T002820	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	<3.010	<3.01e0	n/a	n/a	n/a	3.010	n/a
S98T002820	D	Magnesium-ICP-Acid Dil.	ug/mL	103.8	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T002820	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.6	<5.00e-02	<36.00	<37.0e1	36.55	3.01	n/a	30.10	n/a
S98T002820	D	Nickel-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	<5.32e+02	<5.30e2	536.0	8.63	n/a	60.10	n/a
S98T002820	D	Phosphorus-ICP-Acid Dil.	ug/mL	101.0	<1.00e-01	70.80	70.90	70.85	0.14	n/a	120.0	n/a
S98T002820	D	Selenium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	9.38e+02	9.72e2	955.0	3.56	n/a	60.10	n/a
S98T002820	D	Antimony-ICP-Acid Dil.	ug/mL	101.8	<6.00e-02	<36.10	<36.01e1	n/a	n/a	n/a	36.10	n/a
S98T002820	D	Samarium-ICP-Acid Dil.	ug/mL	97.00	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T002820	D	Strontium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	<6.10	<6.01e1	n/a	n/a	n/a	60.10	n/a
S98T002820	D	Titanium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T002820	D	Thallium-ICP-Acid Dil.	ug/mL	101.0	<1.00e-02	<6.010	<6.01e0	n/a	n/a	n/a	6.010	n/a
S98T002820	D	Vanadium-ICP-Acid Dil.	ug/mL	95.60	<2.00e-01	<12.00e+02	<12.00e2	n/a	n/a	n/a	120.0	n/a
S98T002820	D	Acetate by IC-DIOMEX 4000/4500	ug/mL	99.40	<5.00e-02	<30.10	<30.01e1	n/a	n/a	n/a	30.10	n/a
S98T002820	D	Glycolate by IC-DIOMEX 4500 DRGAC	ug/mL	95.73	<2.00e-02	2.31e+03	2.32e+03	2.32e+03	0.43	n/a	35.15	n/a
S98T002820	D	Cobalt-60 by GEA	uCi/mL	103.3	<1.24e-03	<33.94	<33.94e1	n/a	n/a	n/a	33.94	n/a
S98T002828	D	ILC by Acid/Coulometry	ug/mL	100.5	8.00e-01	3.14e+03	3.23e+03	3.18e+03	2.83	n/a	1.30e-02	n/a
S98T002828	D	Sn 117 by ICP/MS Acid Addition	ug/mL	105.5	<1.50e-04	<6.075	<6.08e0	n/a	n/a	n/a	6.075	n/a
S98T002828	D	Cs 133 by ICP/MS Acid Addition	ug/mL	89.00	<2.00e-06	<2.369	<2.369e1	6.350	0.47	n/a	8.10e-02	n/a
S98T002828	D	Cs 135 by ICP/MS Acid Addition	ug/mL	89.00	<2.00e-06	<2.369	<2.369e1	2.355	3.00	n/a	8.10e-02	n/a
S98T002828	D	Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	<2.00e-06	<2.369	<2.369e1	2.254	1.33	n/a	8.10e-02	n/a
S98T002828	A	Protactinium-231 by ICP/MS	ug/mL	n/a	<5.57e-06	<2.26e-01	<2.26e-1	n/a	n/a	n/a	2.26e-01	n/a
S98T002828	A	Thorium-229 by ICP/MS	ug/mL	n/a	<4.66e-06	<1.81e-01	<1.81e-1	n/a	n/a	n/a	1.81e-01	n/a
S98T002828	A	Thorium-230 by ICP/MS	ug/mL	n/a	<3.55e-06	<1.44e-01	<1.44e-1	n/a	n/a	n/a	1.44e-01	n/a
S98T002828	A	Thorium-232 by ICP/MS	ug/mL	99.20	<1.93e-06	<7.82e-02	<7.82e-2	n/a	n/a	n/a	7.80e-02	n/a
S98T002828	A	Plutonium-241 by ICP/MS	ug/mL	109.7	<3.71e-06	<1.50e-01	<1.50e-1	n/a	n/a	n/a	1.50e-01	n/a
S98T002828	A	Am 243/Cm 243 by ICP/MS	ug/mL	n/a	<3.83e-06	<1.55e-01	<1.55e-1	n/a	n/a	n/a	1.55e-01	n/a

Sample#	R	Ad#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002828	A	Pu	244Cm 244 by ICP/MS	ug/ml	n/a	<3.41e-06	<1.38e-01	<1.38e-1	n/a	n/a	n/a	1.38e-01	n/a

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 2468
SEGMENT #: 9R

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002845	F	Silver - ICP-Fusion	ug/g	97.60	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	201.0	n/a
S981002845	F	Arsenic - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Beryllium - ICP-Fusion	ug/g	101.0	<5.00e-03	<1.01e-02	<1.02e2	n/a	n/a	n/a	100.0	n/a
S981002845	F	Cadmium - ICP-Fusion	ug/g	101.6	<5.00e-03	<1.01e-02	<1.02e2	n/a	n/a	n/a	100.0	n/a
S981002845	F	Cerium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Cobalt - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	402.0	n/a
S981002845	F	Copper - ICP-Fusion	ug/g	106.60	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	201.0	n/a
S981002845	F	Lanthanum - ICP-Fusion	ug/g	102.4	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	1.00e+03	n/a
S981002845	F	Magnesium - ICP-Fusion	ug/g	109.4	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	1.00e+03	n/a
S981002845	F	Molybdenum - ICP-Fusion	ug/g	100.8	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	1.00e+03	n/a
S981002845	F	Neodymium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Phosphorus - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	4.02e+03	n/a
S981002845	F	Lead - ICP-Fusion	ug/g	100.4	<1.00e-01	<2.01e-02	<2.03e2	8.81e+03	6.13	n/a	4.02e+03	n/a
S981002845	F	Sulfur - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Antimony - ICP-Fusion	ug/g	101.4	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Selenium - ICP-Fusion	ug/g	96.40	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	1.21e+03	n/a
S981002845	F	Samarium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Strontium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	2.01e+03	n/a
S981002845	F	Titanium - ICP-Fusion	ug/g	101.8	<1.00e-01	<2.01e-02	<2.03e2	8.24e+03	6.51	n/a	201.0	n/a
S981002845	F	Thallium - ICP-Fusion	ug/g	95.40	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	201.0	n/a
S981002845	F	Vanadium - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	1.00e+03	n/a
S981002845	F	Zinc - ICP-Fusion	ug/g	108.4	<1.00e-01	<2.01e-02	<2.03e2	n/a	n/a	n/a	8.40e+02	n/a
S981002845	F	Cobalt-60 by GEA	uCi/g	111.7	5.00e-03	< 8.331	< 7.652	n/a	n/a	n/a	632.14	n/a
S981002847	A	Cs133 ICP/MS Act Dig LA-505159	ug/g	97.50	2.42e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Cs135 ICP/MS Act Dig LA-505159	ug/g	97.50	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Cs137 ICP/MS Act Dig LA-505159	ug/g	n/a	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Protactinium-231 by ICP/MS	ug/g	n/a	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Thorium-229 by ICP/MS	ug/g	n/a	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Thorium-230 by ICP/MS	ug/g	n/a	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Thorium-232 by ICP/MS	ug/g	n/a	4.62e-05	< 8.331	< 7.652	11.76	28.4	n/a	8.431	n/a
S981002847	A	Pu/Am-241 by ICP/MS	ug/g	102.3	<1.97e-05	< 3.063	< 2.780	n/a	n/a	n/a	2.782	n/a
S981002847	A	Am 243/Am 244 by ICP/MS	ug/g	n/a	<1.97e-05	< 3.063	< 2.780	n/a	n/a	n/a	2.782	n/a
S981002847	A	Pu 244/Am 244 by ICP/MS	ug/g	n/a	<1.97e-05	< 3.063	< 2.780	n/a	n/a	n/a	2.782	n/a
S981002847	A	Silver - ICP-Acid Digest	ug/g	86.20	<1.00e-01	< 6.331	< 5.930	18.35	1.63	n/a	6.331	n/a
S981002847	A	Arsenic - ICP-Acid Digest	ug/g	91.80	<1.00e-01	< 6.331	< 5.930	18.35	1.63	n/a	6.331	n/a
S981002847	A	Beryllium - ICP-Acid Digest	ug/g	103.2	<5.00e-03	< 3.110	< 2.850	n/a	n/a	n/a	3.120	n/a
S981002847	A	Cadmium - ICP-Acid Digest	ug/g	91.80	<5.00e-03	< 3.110	< 2.850	n/a	n/a	n/a	3.120	n/a
S981002847	A	Cerium - ICP-Acid Digest	ug/g	98.20	<1.00e-01	< 6.331	< 5.930	n/a	n/a	n/a	62.40	n/a
S981002847	A	Cobalt - ICP-Acid Digest	ug/g	92.80	<1.00e-02	< 12.50	< 1.13e1	n/a	n/a	n/a	12.50	n/a
S981002847	A	Copper - ICP-Acid Digest	ug/g	90.00	<1.00e-02	< 6.230	< 5.660	n/a	n/a	n/a	6.200	n/a
S981002847	A	Lanthanum - ICP-Acid Digest	ug/g	98.20	<1.00e-02	< 31.10	< 2.85e1	n/a	n/a	n/a	31.20	n/a
S981002847	A	Magnesium - ICP-Acid Digest	ug/g	95.40	<1.00e-01	< 5.95e+02	< 851.0	733.0	35.4	n/a	62.40	n/a
S981002847	A	Molybdenum - ICP-Acid Digest	ug/g	95.40	<1.00e-02	< 31.10	< 2.85e1	n/a	n/a	n/a	31.20	n/a
S981002847	A	Neodymium - ICP-Acid Digest	ug/g	97.40	<1.00e-01	< 64.40	< 67.60	65.00	4.85	n/a	62.40	n/a

5078

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002847	A	Phosphorus - ICP-Acid Digest	ug/g	93.20	<2.00e-01	7.36e+03	1.35e+04	1.04e+04	58.9	N/A	124.0	N/A
S98T002847	A	Lead - ICP-Acid Digest	ug/g	88.80	<1.00e-01	3.22e+02	3.40	331.0	5.44	N/A	62.40	N/A
S98T002847	A	Lead - ICP-Acid Digest	ug/g	87.60	<1.00e-01	1.21e+03	3.40e+03	2.30e+03	95.0	N/A	62.40	N/A
S98T002847	A	Antimony - ICP-Acid Digest	ug/g	91.20	<6.00e-02	< 37.40	<5.40e1	N/A	N/A	N/A	37.40	N/A
S98T002847	A	Selenium - ICP-Acid Digest	ug/g	91.00	<1.00e-01	< 62.30	<5.66e1	N/A	N/A	N/A	62.40	N/A
S98T002847	A	Samarium - ICP-Acid Digest	ug/g	97.20	<1.00e-01	< 62.30	<5.66e1	N/A	N/A	N/A	62.40	N/A
S98T002847	A	Strontium - ICP-Acid Digest	ug/g	97.00	<1.00e-02	7.93e+03	8.39e+03	8.16e+03	5.64	N/A	6.240	N/A
S98T002847	A	Titanium - ICP-Acid Digest	ug/g	97.80	<1.00e-02	51.20	68.70	59.95	29.2	N/A	6.240	N/A
S98T002847	A	Thallium - ICP-Acid Digest	ug/g	88.40	<2.00e-01	<1.25e+02	<1.36e1	N/A	N/A	N/A	124.0	N/A
S98T002847	A	Vanadium - ICP-Acid Digest	ug/g	93.20	<5.00e-02	< 31.70	<2.83e1	N/A	N/A	N/A	84.32	N/A
S98T002847	M	Acetate by IC-DIONEX 4000/4500	ug/g	99.14	2.80e-02	9.06e+02	788.0	847.2	13.9	N/A	84.32	N/A
S98T002849	M	Glycolate IC-DIONEX 4000/4500	ug/g	99.14	<5.60e-02	<2.36e+02	<2.33e2	N/A	N/A	N/A	236.0	N/A

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002821	D	Silver-ICP-Acid Dil.	ug/mL	98.20	<1.00e-02	16.20	15.80	16.00	2.50	N/A	6.010	N/A
S98T002821	D	Arsenic-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Beryllium-ICP-Acid Dil.	ug/mL	101.6	<5.00e-03	< 3.010	<3.01e0	N/A	N/A	N/A	3.000	N/A
S98T002821	D	Cadmium-ICP-Acid Dil.	ug/mL	100.6	<5.00e-03	< 3.010	<3.01e0	N/A	N/A	N/A	3.000	N/A
S98T002821	D	Cerium-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Cobalt-ICP-Acid Dil.	ug/mL	98.20	<2.00e-02	< 12.00	<1.20e1	N/A	N/A	N/A	12.00	N/A
S98T002821	D	Copper-ICP-Acid Dil.	ug/mL	94.60	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002821	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	30.10	N/A
S98T002821	D	Magnesium-ICP-Acid Dil.	ug/mL	102.8	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.2	<5.00e-02	< 30.60	<3.150	31.05	2.90	N/A	30.10	N/A
S98T002821	D	Neodymium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Lead-ICP-Acid Dil.	ug/mL	99.20	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Antimony-ICP-Acid Dil.	ug/mL	96.80	<1.00e-01	< 1.47e+03	1.49e+03	1.48e+03	1.35	N/A	60.10	N/A
S98T002821	D	Selenium-ICP-Acid Dil.	ug/mL	100.6	<6.00e-02	< 36.10	<3.61e1	N/A	N/A	N/A	36.10	N/A
S98T002821	D	Samarium-ICP-Acid Dil.	ug/mL	98.20	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S98T002821	D	Strontium-ICP-Acid Dil.	ug/mL	98.60	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002821	D	Titanium-ICP-Acid Dil.	ug/mL	101.6	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S98T002821	D	Thallium-ICP-Acid Dil.	ug/mL	95.40	<5.00e-01	<1.20e+02	<1.20e2	N/A	N/A	N/A	120.0	N/A
S98T002821	D	Vanadium-ICP-Acid Dil.	ug/mL	99.20	<5.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	30.10	N/A
S98T002821	A	Acetate by IC-DIONEX 4000/4500	ug/mL	95.73	<2.00e-02	2.49e+03	2.60e+03	2.54e+03	4.52	N/A	35.15	N/A
S98T002821	A	Glycolate IC-DIONEX 4500 ORGAC	ug/mL	106.9	<2.00e-02	33.92	<3.39e1	N/A	N/A	N/A	35.15	N/A
S98T002821	A	Cobalt-60 by GEA	ug/mL	103.3	<1.24e+03	<1.32e+2	<1.32e+2	N/A	N/A	N/A	1.10e+02	N/A
S98T002829	TIC by Acid/Countmetry		ug/mL	100.5	8.00e-01	2.72e+03	2.85e+03	2.78e+03	4.61	N/A	1.10e+02	N/A
S98T002829	D	Sn 117 by ICP/MS Acid Addition	ug/mL	111.7	<1.50e-04	< 3.045	<3.05e0	N/A	N/A	N/A	3.045	N/A
S98T002829	D	Cs 133 by ICP/MS Acid Addition	ug/mL	98.50	<2.00e-06	6.705	9.770	6.568	3.64	N/A	4.10e-02	N/A
S98T002829	D	Cs 137 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	3.453	2.440	2.449	0.82	N/A	4.10e-02	N/A
S98T002829	A	Protactinium-231 by ICP/MS	ug/mL	N/A	<5.75e-06	<2.62e-01	<2.62e-1	N/A	N/A	N/A	2.20e-02	N/A
S98T002829	A	Thorium-229 by ICP/MS	ug/mL	N/A	<4.74e-06	<1.81e-01	<1.81e-1	N/A	N/A	N/A	1.81e-01	N/A
S98T002829	A	Thorium-230 by ICP/MS	ug/mL	N/A	<3.55e-06	<1.44e-01	<1.44e-1	N/A	N/A	N/A	1.44e-01	N/A
S98T002829	A	Thorium-232 by ICP/MS	ug/mL	99.70	<3.72e-06	<1.86e-02	<1.86e-2	N/A	N/A	N/A	7.80e-02	N/A
S98T002829	A	Pu/Am-241 by ICP/MS	ug/mL	109.7	<3.72e-06	<1.50e-01	<1.50e-1	N/A	N/A	N/A	1.50e-01	N/A
S98T002829	A	Am 243/Am 243 by ICP/MS	ug/mL	N/A	<3.33e-06	<1.35e-01	<1.35e-1	N/A	N/A	N/A	1.55e-01	N/A

25-mar-1999 10:30:17
A-0002-1

Sample#	R	A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98100829	A	PU	244/Em 244 by ICP/MS	ug/mL	n/a	<3.41e-06	<1.38e-01	<1.38e-1	n/a	n/a	n/a	1.38e-01	n/a	n/a

5080

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 2468

SEGMENT #: Core Composite

SEGMENT PORTION: Core Composite

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T003364	D	Silver-ICP-Acid Dil.	ug/mL	98.40	< 1.00e-02	15.20	15.40	15.30	1.31	91.70	6.010	N/A
S98T003364	D	Aluminum-ICP-Acid Dil.	ug/mL	99.60	< 1.00e-01	< 9.010	< 9.010	N/A	N/A	98.50	60.10	N/A
S98T003364	D	Beryllium-ICP-Acid Dil.	ug/mL	101.8	< 5.00e-03	3.010	< 3.010	N/A	N/A	98.70	3.000	N/A
S98T003364	D	Cadmium-ICP-Acid Dil.	ug/mL	100.8	< 5.00e-03	2.010	< 2.010	N/A	N/A	99.40	3.000	N/A
S98T003364	D	Cerium-ICP-Acid Dil.	ug/mL	100.8	< 1.00e-01	< 6.010	< 6.010	N/A	N/A	99.20	60.10	N/A
S98T003364	D	Cobalt-ICP-Acid Dil.	ug/mL	99.60	< 2.00e-02	12.00	< 12.00	N/A	N/A	97.50	12.00	N/A
S98T003364	D	Copper-ICP-Acid Dil.	ug/mL	96.40	< 1.00e-02	< 5.010	< 5.010	N/A	N/A	97.70	60.10	N/A
S98T003364	D	Lead-ICP-Acid Dil.	ug/mL	103.8	< 5.00e-02	< 6.010	< 6.010	N/A	N/A	97.30	30.10	N/A
S98T003364	D	Mercury-ICP-Acid Dil.	ug/mL	104.8	< 1.00e-01	< 40.10	< 40.10	N/A	N/A	98.30	60.10	N/A
S98T003364	D	Molybdenum-ICP-Acid Dil.	ug/mL	98.40	< 5.00e-02	33.70	34.10	33.90	1.18	94.90	30.10	N/A
S98T003364	D	Nickel-ICP-Acid Dil.	ug/mL	98.40	< 1.00e-01	< 60.10	< 60.10	N/A	N/A	94.90	60.10	N/A
S98T003364	D	Phosphorus-ICP-Acid Dil.	ug/mL	105.6	< 2.00e-01	4.00e+02	408.0	407.0	0.49	104.0	120.0	N/A
S98T003364	D	Lead-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-01	69.20	62.10	65.65	10.8	99.90	60.10	N/A
S98T003364	D	Sulfur-ICP-Acid Dil.	ug/mL	101.2	< 1.00e-01	9.21e+02	935.0	928.0	1.51	101.0	60.10	N/A
S98T003364	D	Antimony-ICP-Acid Dil.	ug/mL	106.0	< 5.00e-02	< 36.10	< 36.10	N/A	N/A	104.0	36.10	N/A
S98T003364	D	Samarium-ICP-Acid Dil.	ug/mL	97.20	< 1.00e-01	< 60.10	< 60.10	N/A	N/A	110.0	60.10	N/A
S98T003364	D	Selenium-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-01	< 60.10	< 60.10	N/A	N/A	94.60	60.10	N/A
S98T003364	D	Strontium-ICP-Acid Dil.	ug/mL	106.2	< 1.00e-02	< 6.010	< 6.010	N/A	N/A	103.0	6.010	N/A
S98T003364	D	Titanium-ICP-Acid Dil.	ug/mL	95.00	< 2.00e-01	< 1.20e+02	< 1.20e+02	N/A	N/A	92.70	120.0	N/A
S98T003364	D	Thallium-ICP-Acid Dil.	ug/mL	100.0	< 5.00e-02	< 30.10	< 30.10	N/A	N/A	97.60	30.10	N/A
S98T003364	D	Vanadium-ICP-Acid Dil.	ug/mL	91.94	< 2.00e-02	1.41e+03	1.54e+03	1.47e+03	8.81	97.58	298.8	N/A
S98T003364	D	Acetate by IC-Dionex 4000/4500	ug/mL	102.8	< 2.00e-02	< 2.88e+02	< 2.88e+02	N/A	N/A	104.7	288.8	N/A
S98T003364	D	Glycolate by IC-Dionex 4500 ORGAC	ug/mL	106.8	< 7.93e-04	< 3.74e+03	< 3.93e+3	N/A	N/A	N/A	4.00e-03	N/A
S98T003364	D	Cobalt-60 by GEA	uCi/mL	99.00	< 1.700	4.52e+03	4.48e+03	4.50e+03	1.79	99.40	5.000	N/A
S98T003365	D	TIC by GC/Colometry	ug/mL	105.5	< 1.50e-04	< 6.075	< 6.080	N/A	N/A	89.64	6.075	N/A
S98T003365	D	Cs 117 by ICP/MS Acid Addition	ug/mL	89.00	< 2.00e-06	5.921	5.760	5.841	2.74	89.55	8.10e-02	N/A
S98T003365	D	Cs 133 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	2.090	2.070	2.080	0.96	N/A	8.10e-02	N/A
S98T003365	D	Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	2.021	1.980	2.000	2.00	N/A	8.10e-02	N/A
S98T003365	D	Protactinium-231 by ICP/MS	ug/mL	n/a	< 5.57e-06	< 2.26e-01	< 2.26e-1	N/A	N/A	N/A	1.81e-01	N/A
S98T003365	D	Thorium-229 by ICP/MS	ug/mL	n/a	< 4.46e-06	< 1.81e-01	< 1.81e-1	N/A	N/A	N/A	1.44e-01	N/A
S98T003365	D	Thorium-230 by ICP/MS	ug/mL	n/a	< 5.59e-06	< 1.44e-01	< 1.44e-1	N/A	N/A	N/A	7.80e-02	N/A
S98T003365	D	Thorium-232 by ICP/MS	ug/mL	99.20	< 1.93e-06	< 7.82e-02	< 7.82e-2	N/A	N/A	94.00	7.80e-02	N/A
S98T003365	D	Pu Am-241 by ICP/MS	ug/mL	109.7	< 3.71e-06	< 1.50e-01	< 1.50e-1	N/A	N/A	105.3	1.50e-01	N/A
S98T003365	D	Pu Am-241 by ICP/MS	ug/mL	n/a	< 8.83e-06	< 1.58e-01	< 1.58e-1	N/A	N/A	N/A	1.58e-01	N/A
S98T003365	D	Pu 243/244 by ICP/MS	ug/mL	n/a	< 3.41e-06	< 1.38e-01	< 1.38e-1	N/A	N/A	N/A	1.38e-01	N/A
S98T003365	D	Pu 244/244 by ICP/MS	ug/mL	n/a	< 3.41e-06	< 1.38e-01	< 1.38e-1	N/A	N/A	N/A	1.38e-01	N/A

Appendix A. Opportunistic Analyte Results.
87-105 (2)

CORE NUMBER: 250
SEGMENT #: 3

SEGMENT PORTION: L Lower Half of Segment

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002739	F	Silver - ICP-Fusion	ug/g	100.4	< 1.00e-01	< 2.00e-02	< 2.04e-2	n/a	n/a	81.40	205.0	n/a
S981002739	F	Antimony - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 2.00e-03	< 2.04e-3	n/a	n/a	94.70	2.05e+03	n/a
S981002739	F	Barium - ICP-Fusion	ug/g	103.4	< 5.00e-03	< 1.02e+02	< 1.02e-2	n/a	n/a	93.00	102.0	n/a
S981002739	F	Cadmium - ICP-Fusion	ug/g	103.4	< 5.00e-03	< 1.02e+02	< 1.02e-2	n/a	n/a	93.00	102.0	n/a
S981002739	F	Cerium - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 4.00e+03	< 4.07e-2	n/a	n/a	85.60	2.05e+03	n/a
S981002739	F	Cobalt - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 4.00e+02	< 4.07e-2	n/a	n/a	94.40	409.0	n/a
S981002739	F	Copper - ICP-Fusion	ug/g	98.6	< 1.00e-02	< 1.02e+02	< 2.04e-2	n/a	n/a	85.90	205.0	n/a
S981002739	F	Lanthanum - ICP-Fusion	ug/g	105.8	< 1.00e-02	< 1.02e+03	< 1.02e-3	n/a	n/a	90.10	1.02e+03	n/a
S981002739	F	Magnesium - ICP-Fusion	ug/g	105.8	< 1.00e-01	< 1.02e+03	< 1.02e-3	n/a	n/a	90.60	1.02e+03	n/a
S981002739	F	Molybdenum - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 1.02e+03	< 1.02e-3	n/a	n/a	95.90	1.02e+03	n/a
S981002739	F	Neodymium - ICP-Fusion	ug/g	100.2	< 1.00e-01	< 1.02e+03	< 1.02e-3	n/a	n/a	89.70	2.05e+03	n/a
S981002739	F	Phosphorus - ICP-Fusion	ug/g	103.0	< 2.00e-01	< 5.00e+03	< 5.04e-3	5.70e+03	n/a	94.10	4.09e+03	n/a
S981002739	F	Lead - ICP-Fusion	ug/g	102.2	< 1.00e-01	< 2.00e+03	< 2.04e-3	n/a	n/a	100.0	2.05e+03	n/a
S981002739	F	Sulfur - ICP-Fusion	ug/g	98.6	< 1.00e-02	< 1.02e+03	< 1.02e-3	n/a	n/a	102.0	2.05e+03	n/a
S981002739	F	Antimony - ICP-Fusion	ug/g	102.8	< 1.00e-02	< 1.02e+03	< 1.02e-3	n/a	n/a	100.0	2.05e+03	n/a
S981002739	F	Selenium - ICP-Fusion	ug/g	98.20	< 1.00e-01	< 2.00e+03	< 2.04e-3	n/a	n/a	95.40	1.23e+03	n/a
S981002739	F	Samarium - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.00e+03	< 2.04e-3	n/a	n/a	93.50	2.05e+03	n/a
S981002739	F	Strontium - ICP-Fusion	ug/g	100.0	< 1.00e-02	< 2.00e+03	< 2.04e-3	n/a	n/a	91.20	2.05e+03	n/a
S981002739	F	Titanium - ICP-Fusion	ug/g	103.2	< 1.00e-02	< 2.00e+03	< 2.04e-3	n/a	n/a	88.70	205.0	n/a
S981002739	F	Thallium - ICP-Fusion	ug/g	96.80	< 1.00e-01	< 4.00e+03	< 4.07e-2	n/a	n/a	92.30	205.0	n/a
S981002739	F	Vanadium - ICP-Fusion	ug/g	101.0	< 5.00e-02	< 1.02e+03	< 1.02e-3	n/a	n/a	92.30	4.09e+03	n/a
S981002739	F	Vanadium - ICP-Fusion	ug/g	104.3	< 4.14e-02	< 1.01e+02	< 1.01e-2	n/a	n/a	93.60	1.00e+02	n/a
S981002739	F	Cobalt-60 by GEA	uCi/g	106.3	7.00e-03	< 6.15e	< 6.85e-3	n/a	n/a	78.20	6.15e	n/a
S981002740	A	Sn117 ICP/MS Acid Dig LA-505159	ug/g	99.00	2.75e-05	8.02e-01	9.16e-01	8.62e-01	12.5	100.0	8.20e-02	n/a
S981002740	A	Cs133 ICP/MS Acid Dig LA-505159	ug/g	n/a	6.93e-05	4.72e-01	4.37e-01	4.37e-01	9.52	n/a	8.20e-02	n/a
S981002740	A	Cs137 ICP/MS Acid Dig LA-505159	ug/g	n/a	1.11e-04	1.100	1.210	1.155	n/a	n/a	1.63e-02	n/a
S981002740	A	Protactinium-231 by ICP/MS	ug/g	n/a	< 4.36e-05	< 1.63e-01	< 1.54e-1	n/a	n/a	n/a	2.24e-01	n/a
S981002740	A	Thorium-229 by ICP/MS	ug/g	n/a	< 6.00e-05	< 2.24e-01	< 2.13e-1	n/a	n/a	n/a	2.24e-01	n/a
S981002740	A	Thorium-230 by ICP/MS	ug/g	n/a	< 1.84e-05	< 6.86e-02	< 6.51e-2	n/a	n/a	n/a	6.90e-02	n/a
S981002740	A	Pu/Am-241 by ICP/MS	ug/g	107.0	< 4.12e-05	5.654	5.350	5.502	5.45	n/a	1.54e-01	n/a
S981002740	A	Pu/Am-241 by ICP/MS	ug/g	103.9	< 1.85e-05	< 6.88e-02	< 6.54e-2	n/a	n/a	n/a	6.90e-02	n/a
S981002740	A	Pu 243/Am 244 by ICP/MS	ug/g	n/a	< 3.68e-02	< 3.20e-1	< 3.20e-1	n/a	n/a	n/a	3.70e-02	n/a
S981002740	A	Pu 243/Am 244 by ICP/MS	ug/g	n/a	< 6.31e-04	< 2.35e-01	< 2.24e-1	n/a	n/a	n/a	2.35e-01	n/a
S981002740	A	Silver - ICP-Acid Digest	ug/g	86.30	< 1.00e-02	16.40	17.30	16.85	5.34	95.50	10.20	n/a
S981002740	A	Arsenic - ICP-Acid Digest	ug/g	93.60	< 1.00e-01	< 1.02e+02	< 9.70e1	n/a	n/a	104.2	102.0	n/a
S981002740	A	Beryllium - ICP-Acid Digest	ug/g	93.20	< 5.00e-03	< 5.100	< 4.85e0	n/a	n/a	102.2	5.100	n/a
S981002740	A	Cadmium - ICP-Acid Digest	ug/g	93.20	< 5.00e-03	< 5.100	< 5.77e0	n/a	n/a	102.2	5.100	n/a
S981002740	A	Cerium - ICP-Acid Digest	ug/g	97.20	< 1.00e-01	< 1.02e+02	< 9.70e1	n/a	n/a	99.60	102.0	n/a
S981002740	A	Cobalt - ICP-Acid Digest	ug/g	94.00	< 2.00e-02	< 20.40	< 1.94e1	n/a	n/a	100.6	20.40	n/a
S981002740	A	Copper - ICP-Acid Digest	ug/g	88.60	< 1.00e-02	< 10.20	< 9.70e0	n/a	n/a	92.60	10.20	n/a
S981002740	A	Lanthanum - ICP-Acid Digest	ug/g	97.00	< 5.00e-02	< 5.100	< 4.85e1	n/a	n/a	100.2	51.00	n/a
S981002740	A	Magnesium - ICP-Acid Digest	ug/g	99.00	< 1.00e-01	< 1.02e+02	< 9.70e1	n/a	n/a	105.6	102.0	n/a
S981002740	A	Molybdenum - ICP-Acid Digest	ug/g	99.00	< 5.00e-02	< 5.100	< 4.85e1	n/a	n/a	102.0	51.00	n/a
S981002740	A	Neodymium - ICP-Acid Digest	ug/g	96.00	< 1.00e-01	< 1.02e+02	< 9.70e1	n/a	n/a	98.40	102.0	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002740	A	Phosphorus -ICP-Acid Digest	ug/g	92.60	<2.00e-01	6.56e-03	5.67e-03	6.19e-03	11.3	264.0	200.0	0/A
S981002740	A	Lead -ICP-Acid Digest	ug/g	99.60	<1.00e-01	1.02e-02	8.67e-03	n/a	n/a	100.0	100.0	0/A
S981002740	A	Sulfur -ICP-Acid Digest	ug/g	99.40	1.50e-01	1.61e-02	1.85e-02	1.73e-02	13.9	100.0	100.0	0/A
S981002740	A	Antimony -ICP-Acid Digest	ug/g	91.80	<5.00e-02	8.12e-03	8.82e-03	n/a	n/a	100.0	100.0	0/A
S981002740	A	Selenium -ICP-Acid Digest	ug/g	92.80	<1.00e-01	4.05e-02	<9.70e-03	n/a	n/a	100.0	100.0	0/A
S981002740	A	Strontium -ICP-Acid Digest	ug/g	96.80	<1.00e-01	4.10e-02	<9.70e-03	n/a	n/a	97.46	100.0	0/A
S981002740	A	Vanadium -ICP-Acid Digest	ug/g	96.80	<1.00e-02	4.72e-02	4.72e-02	31.10	12.9	98.90	100.0	0/A
S981002740	A	Thallium -ICP-Acid Digest	ug/g	99.80	<2.00e-02	2.00e-02	4.94e-02	n/a	n/a	100.0	100.0	0/A
S981002740	A	Vanadium -ICP-Acid Digest	ug/g	99.80	<5.00e-01	5.11e-02	44.85e-02	n/a	n/a	100.0	51.00	0/A
S981002740	M	Verdate by IC-DIONEY 4000/4500	ug/g	100.9	3.20e-02	4.81e-02	44.89e-02	474.9	2.55	95.49	84.13	0/A
S981002741	M	Verdate by IC-DIONEY 4000/4500	ug/g	93.10	<5.40e-02	<2.36e-02	<2.38e-02	n/a	n/a	96.55	239.0	0/A

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 250
SEGMENT #: 4

SEGMENT PORTION: L Lower Half of Segment

Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPB % Spk Rec %	Det Limit	Count	Err%
S98T002918	F	Silver - ICP-Fusion	ug/g	99.60	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	20.0	N/A
S98T002918	F	Arsenic - ICP-Fusion	ug/g	100.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Beryllium - ICP-Fusion	ug/g	101.0	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	100.0	N/A
S98T002918	F	Cadmium - ICP-Fusion	ug/g	101.0	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Cerium - ICP-Fusion	ug/g	100.0	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Cobalt - ICP-Fusion	ug/g	99.60	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	200.0	N/A
S98T002918	F	Copper - ICP-Fusion	ug/g	100.60	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.00e+03	N/A
S98T002918	F	Lead - ICP-Fusion	ug/g	100.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Magnesium - ICP-Fusion	ug/g	100.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.00e+03	N/A
S98T002918	F	Molybdenum - ICP-Fusion	ug/g	100.8	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Nickel - ICP-Fusion	ug/g	99.60	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Phosphorus - ICP-Fusion	ug/g	100.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	101.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	200.0	N/A
S98T002918	F	Silver - ICP-Fusion	ug/g	101.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	4.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	96.40	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	97.20	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	98.00	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	101.8	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	95.40	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	99.80	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	108.4	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	2.00e+03	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	111.7	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	5.970	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	97.50	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	8.00e-02	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	N/A	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	8.00e-02	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	N/A	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	8.00e-02	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	N/A	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.31e-01	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	N/A	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.28e-01	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	102.0	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.28e-01	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	102.3	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	7.10e-02	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	N/A	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	1.03e-01	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	86.20	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	9.900	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	91.80	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	4.950	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	103.2	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	4.950	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	98.20	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	92.00	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	92.80	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	19.80	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	90.00	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	9.900	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	98.20	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	49.50	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	95.40	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	92.00	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	95.60	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	49.50	N/A
S98T002918	F	Selenium - ICP-Fusion	ug/g	97.40	< 1.00e-03	< 2.00e-02	< 2.00e-02	N/A	N/A	N/A	92.00	N/A

Sample #	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002926	A	Phosphorus -ICP-Acid Digest	ug/g	93.20	<2.00e-01	1.86e+03	5.45e+03	3.64e+03	95.9	N/A	198.0	N/A
S981002926	A	Lead -ICP-Acid Digest	ug/g	88.80	<1.00e-01	< 97.00	<1.00e2	N/A	N/A	N/A	97.00	N/A
S981002926	A	Sulfur -ICP-Acid Digest	ug/g	87.60	<1.00e-01	1.19e+03	1.30e+03	1.24e+03	8.84	N/A	97.00	N/A
S981002926	A	Antimony -ICP-Acid Digest	ug/g	91.20	<6.00e-02	< 97.40	<5.01e1	N/A	N/A	N/A	97.20	N/A
S981002926	A	Selenium -ICP-Acid Digest	ug/g	91.00	<1.00e-01	< 97.00	<1.00e2	N/A	N/A	N/A	97.00	N/A
S981002926	A	Samarium -ICP-Acid Digest	ug/g	97.20	<1.00e-01	< 97.00	<1.00e2	N/A	N/A	N/A	97.00	N/A
S981002926	A	Strontium -ICP-Acid Digest	ug/g	97.00	<1.00e-02	< 17.60	<14.30	15.35	20.7	N/A	97.00	N/A
S981002926	A	Titanium-ICP-Acid Digest	ug/g	97.80	<1.00e-02	< 9.200	<1.00e1	N/A	N/A	N/A	97.00	N/A
S981002926	A	Thallium -ICP-Acid Digest	ug/g	88.40	<2.00e-01	<1.98e+02	<2.00e2	N/A	N/A	N/A	198.0	N/A
S981002926	A	Vanadium -ICP-Acid Digest	ug/g	93.20	<5.00e-02	< 47.50	<3.91e1	N/A	N/A	N/A	49.0	N/A
S981002931	W	Acetate by IC-DIONEX 4000/4500	ug/g	17.48	<2.00e-02	1.79e+02	186.0	182.6	3.84	76.1	83.28	N/A
S981002931	W	Glycolate IC-DIONEX 4000/4500	ug/g	93.10	<3.60e-02	<2.34e+02	<2.29e2	N/A	N/A	91.38	233.7	N/A

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 250
SEGMENT #: 5

SEGMENT PORTION: L Lower Half of Segment

Sample #	B #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count	Err%
S98T002919	F	Silver - ICP-Fusion	ug/g	99.40	< 1.00e-02	< 2.05e-02	< 0.8e-3	n/a	n/a	90.40	205.0	n/a	n/a
S98T002919	F	Aluminum - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.05e-02	< 0.8e-3	n/a	n/a	100.0	2.05e+03	n/a	n/a
S98T002919	F	Beryllium - ICP-Fusion	ug/g	101.0	< 5.00e-03	< 4.03e-02	< 0.6e-3	n/a	n/a	97.30	103.0	n/a	n/a
S98T002919	F	Cadmium - ICP-Fusion	ug/g	101.0	< 5.00e-03	< 4.03e-02	< 0.6e-3	n/a	n/a	101.0	103.0	n/a	n/a
S98T002919	F	Cerium - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.05e-02	< 0.8e-3	n/a	n/a	95.90	2.05e+03	n/a	n/a
S98T002919	F	Cobalt - ICP-Fusion	ug/g	99.20	< 2.00e-02	< 4.11e-02	< 4.0e-3	n/a	n/a	98.30	410.0	n/a	n/a
S98T002919	F	Copper - ICP-Fusion	ug/g	99.40	< 1.00e-02	< 2.05e-02	< 0.8e-3	n/a	n/a	90.30	205.0	n/a	n/a
S98T002919	F	Lanthanum - ICP-Fusion	ug/g	100.4	< 5.00e-02	< 1.03e+03	< 1.6e-3	n/a	n/a	95.60	1.03e+03	n/a	n/a
S98T002919	F	Magnesium - ICP-Fusion	ug/g	102.4	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	98.60	2.05e+03	n/a	n/a
S98T002919	F	Nickel - ICP-Fusion	ug/g	100.8	< 5.00e-02	< 1.03e+03	< 0.8e-3	n/a	n/a	100.0	1.03e+03	n/a	n/a
S98T002919	F	Phosphorus - ICP-Fusion	ug/g	98.00	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	92.60	2.05e+03	n/a	n/a
S98T002919	F	Platinum - ICP-Fusion	ug/g	99.80	< 2.00e-01	< 4.11e+03	< 1.6e-3	n/a	n/a	101.0	4.11e+03	n/a	n/a
S98T002919	F	Selenium - ICP-Fusion	ug/g	100.4	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	99.90	2.05e+03	n/a	n/a
S98T002919	F	Silver - ICP-Fusion	ug/g	98.40	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	103.0	2.05e+03	n/a	n/a
S98T002919	F	Strontium - ICP-Fusion	ug/g	101.4	< 5.00e-02	< 1.23e+03	< 1.2e-3	n/a	n/a	100.0	1.23e+03	n/a	n/a
S98T002919	F	Titanium - ICP-Fusion	ug/g	96.40	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	97.70	2.05e+03	n/a	n/a
S98T002919	F	Tungsten - ICP-Fusion	ug/g	97.20	< 1.00e-01	< 2.05e+03	< 0.8e-3	n/a	n/a	93.00	2.05e+03	n/a	n/a
S98T002919	F	Vanadium - ICP-Fusion	ug/g	95.40	< 1.00e-02	< 2.05e+02	< 0.8e-3	n/a	n/a	98.70	205.0	n/a	n/a
S98T002919	F	Zinc - ICP-Fusion	ug/g	101.8	< 1.00e-02	< 4.11e+03	< 1.6e-3	n/a	n/a	91.70	4.11e+03	n/a	n/a
S98T002919	F	Antimony - ICP-Fusion	ug/g	99.80	< 5.00e-02	< 1.03e+03	< 1.6e-3	n/a	n/a	98.00	1.03e+03	n/a	n/a
S98T002919	F	Barium - ICP-Fusion	ug/g	108.4	< 5.00e-02	< 4.85e+03	< 8.29e-3	n/a	n/a	98.45	8.00e-03	n/a	n/a
S98T002919	F	Bismuth - ICP-Fusion	ug/g	111.7	< 5.00e-03	< 5.907	< 5.907	n/a	n/a	98.45	5.907	n/a	n/a
S98T002927	A	Ca117 ICP/MS Acid Dig LA-505159	ug/g	97.50	2.42e-03	4.69e-01	4.53e-01	7.06	100.5	7.06	2.90e-02	n/a	n/a
S98T002927	A	Ca133 ICP/MS Acid Dig LA-505159	ug/g	97.50	4.62e-03	3.03e-01	2.11e-01	2.57e-01	35.8	n/a	7.90e-02	n/a	n/a
S98T002927	A	Ca137 ICP/MS Acid Dig LA-505159	ug/g	n/a	7.92e-03	7.01e-01	5.32e-01	6.17e-01	27.4	n/a	7.90e-02	n/a	n/a
S98T002927	A	Protactinium-231 by ICP/MS	ug/g	n/a	< 3.63e-03	< 1.30e-01	< 1.26e-1	n/a	n/a	n/a	1.30e-01	n/a	n/a
S98T002927	A	Thorium-229 by ICP/MS	ug/g	n/a	< 4.77e-03	< 1.71e-01	< 1.66e-1	n/a	n/a	n/a	1.71e-01	n/a	n/a
S98T002927	A	Thorium-230 by ICP/MS	ug/g	n/a	< 3.58e-03	< 1.27e-01	< 1.23e-1	n/a	n/a	n/a	1.27e-01	n/a	n/a
S98T002927	A	Thorium-232 by ICP/MS	ug/g	102.3	< 1.97e-03	< 5.87e-01	< 5.69e-01	5.78e-01	3.11	100.9	1.00e-01	n/a	n/a
S98T002927	A	Pu/Am-241 by ICP/MS	ug/g	102.3	< 1.97e-03	< 5.87e-01	< 5.69e-01	5.78e-01	3.11	100.9	1.00e-01	n/a	n/a
S98T002927	A	Pu/Am-243 by ICP/MS	ug/g	n/a	< 2.86e-03	< 1.02e-01	< 9.95e-2	n/a	n/a	n/a	1.02e-01	n/a	n/a
S98T002927	A	Pu/Am-244 by ICP/MS	ug/g	n/a	< 4.20e-03	< 1.50e-01	< 1.46e-1	n/a	n/a	n/a	1.50e-01	n/a	n/a
S98T002927	A	Silver - ICP-Acid Digest	ug/g	86.20	< 1.00e-02	< 98.00	< 9.53e-1	16.90	3.55	95.80	9.800	n/a	n/a
S98T002927	A	Arsenic - ICP-Acid Digest	ug/g	91.80	< 1.00e-02	< 98.00	< 9.53e-1	16.90	3.55	102.4	98.00	n/a	n/a
S98T002927	A	Beryllium - ICP-Acid Digest	ug/g	103.2	< 5.00e-03	< 4.900	< 4.76e0	n/a	n/a	101.4	4.900	n/a	n/a
S98T002927	A	Cadmium - ICP-Acid Digest	ug/g	91.80	< 5.00e-03	< 4.900	< 4.76e0	n/a	n/a	102.2	4.900	n/a	n/a
S98T002927	A	Cerium - ICP-Acid Digest	ug/g	98.20	< 1.00e-01	< 98.00	< 9.53e-1	n/a	n/a	98.80	98.00	n/a	n/a
S98T002927	A	Cobalt - ICP-Acid Digest	ug/g	92.80	< 1.00e-02	< 19.60	< 1.91e-1	n/a	n/a	100.2	19.60	n/a	n/a
S98T002927	A	Copper - ICP-Acid Digest	ug/g	90.00	< 1.00e-02	< 9.800	< 5.53e0	n/a	n/a	97.40	9.800	n/a	n/a
S98T002927	A	Lanthanum - ICP-Acid Digest	ug/g	98.20	< 5.00e-02	< 49.00	< 4.76e1	n/a	n/a	97.80	49.00	n/a	n/a
S98T002927	A	Magnesium - ICP-Acid Digest	ug/g	95.40	< 1.00e-01	< 98.00	< 9.53e-1	n/a	n/a	103.6	98.00	n/a	n/a
S98T002927	A	Nickel - ICP-Acid Digest	ug/g	95.40	< 5.00e-02	< 49.00	< 4.76e1	n/a	n/a	102.4	49.00	n/a	n/a
S98T002927	A	Neodymium - ICP-Acid Digest	ug/g	97.40	< 1.00e-01	< 98.00	< 9.53e-1	n/a	n/a	95.60	98.00	n/a	n/a

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count Err%
S981002927	A	Phosphorus -ICP-Acid Digest	ug/g	99.20	<2.00e-01	1.30e-03	1.25e-03	1.28e-03	3.12	100.8	194.0	0.7
S981002957	A	Lead -ICP-Acid Digest	ug/g	99.20	<1.00e-01	9.8e-05	9.55e-05	9.67e-05	0.4	100.2	98.00	0.7
S981002957	A	Sulfur -ICP-Acid Digest	ug/g	99.20	<1.00e-01	2.97e-03	3.27e-03	3.07e-03	0.4	102.4	98.00	0.7
S981002957	A	Antimony -ICP-Acid Digest	ug/g	99.20	<6.00e-02	5.8e-05	5.72e-05	5.76e-05	0.4	99.4	98.00	0.7
S981002957	A	Selenium -ICP-Acid Digest	ug/g	99.20	<1.00e-01	9.8e-05	9.55e-05	9.67e-05	0.4	99.4	98.00	0.7
S981002957	A	Samarium -ICP-Acid Digest	ug/g	99.20	<1.00e-01	9.8e-05	9.55e-05	9.67e-05	0.4	99.4	98.00	0.7
S981002957	A	Strontium -ICP-Acid Digest	ug/g	99.20	<1.00e-02	9.800e-03	9.550e-03	9.675e-03	0.4	97.40	9.800	0.7
S981002957	A	Thallium -ICP-Acid Digest	ug/g	99.20	<2.00e-01	4.97e-02	4.79e-02	4.88e-02	0.4	95.40	194.0	0.7
S981002957	A	Thallium -ICP-Acid Digest	ug/g	99.20	<2.00e-02	8.9e-02	8.74e-02	8.82e-02	0.4	98.40	49.00	0.7
S981002957	A	Vanadium -ICP-Acid Digest	ug/g	99.20	<2.00e-02	8.9e-02	8.74e-02	8.82e-02	0.4	98.40	83.94	0.7
S981002932	W	Acetate by IC-DIONE	ug/g	79.58	<2.00e-02	83.92	84.51	84.21	0.4	84.03	233.0	0.7
S981002932	W	Glycolate IC-DIONE	4000/4500 ug/g	99.14	<3.00e-02	2.35e-02	2.38e-02	2.36e-02	0.4	98.28	233.0	0.7

CORE NUMBER: 250
SEGMENT #: 6

Appendix A. Opportunistic Analyte Results.
81-105 (2)

SEGMENT PORTION: L Lower Half of Segment

Sample #	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk	Rec %	Det Limit	Count	Err%
S981002920	F	Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	2.01e-02	2.01e-02	n/a	n/a	90.10	90.10	2.01e-03	201.0	n/a
S981002920	F	Arsenic - ICP-Fusion	ug/g	101.3	<1.00e-03	4.01e-03	4.01e-03	n/a	n/a	101.0	101.0	2.01e+03	n/a	n/a
S981002920	F	Beryllium - ICP-Fusion	ug/g	101.3	<1.00e-03	1.00e-02	1.00e-02	n/a	n/a	98.70	98.70	100.0	100.0	n/a
S981002920	F	Barium - ICP-Fusion	ug/g	103.0	<1.00e-03	1.00e-02	1.00e-02	n/a	n/a	102.0	102.0	100.0	100.0	n/a
S981002920	F	Cerium - ICP-Fusion	ug/g	98.60	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	95.40	95.40	2.01e+03	n/a	n/a
S981002920	F	Cobalt - ICP-Fusion	ug/g	100.3	<1.00e-02	4.01e-03	4.01e-03	n/a	n/a	98.70	98.70	402.0	402.0	n/a
S981002920	F	Copper - ICP-Fusion	ug/g	97.40	<1.00e-02	2.01e-02	2.01e-02	n/a	n/a	96.50	96.50	2.01e+03	n/a	n/a
S981002920	F	Lanthanum - ICP-Fusion	ug/g	99.00	<1.00e-02	1.00e-03	1.00e-03	n/a	n/a	99.00	99.00	2.01e+03	n/a	n/a
S981002920	F	Magnesium - ICP-Fusion	ug/g	102.2	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	100.0	100.0	1.00e+03	n/a	n/a
S981002920	F	Nickel - ICP-Fusion	ug/g	96.00	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	93.60	93.60	2.01e+03	n/a	n/a
S981002920	F	Polytheneum - ICP-Fusion	ug/g	101.3	<1.00e-01	4.01e-03	4.01e-03	n/a	n/a	102.0	102.0	4.02e+03	n/a	n/a
S981002920	F	Phosphorus - ICP-Fusion	ug/g	101.3	<1.00e-01	2.01e-03	2.01e-03	n/a	n/a	98.70	98.70	2.01e+03	n/a	n/a
S981002920	F	Lead - ICP-Fusion	ug/g	98.90	<1.00e-01	2.01e-03	2.01e-03	n/a	n/a	105.0	105.0	1.21e+03	n/a	n/a
S981002920	F	Lead - ICP-Fusion	ug/g	101.3	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	96.60	96.60	2.01e+03	n/a	n/a
S981002920	F	Antimony - ICP-Fusion	ug/g	97.00	<1.00e-01	2.01e-03	2.01e-03	n/a	n/a	94.10	94.10	201.0	201.0	n/a
S981002920	F	Selenium - ICP-Fusion	ug/g	97.00	<1.00e-01	2.01e-03	2.01e-03	n/a	n/a	98.70	98.70	2.01e+03	n/a	n/a
S981002920	F	Samarium - ICP-Fusion	ug/g	97.00	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	93.40	93.40	2.01e+03	n/a	n/a
S981002920	F	Strontium - ICP-Fusion	ug/g	100.8	<1.00e-02	2.01e-03	2.01e-03	n/a	n/a	94.10	94.10	201.0	201.0	n/a
S981002920	F	Tantalum - ICP-Fusion	ug/g	96.00	<2.00e-01	4.01e-03	4.01e-03	n/a	n/a	94.20	94.20	4.02e+03	n/a	n/a
S981002920	F	Thallium - ICP-Fusion	ug/g	99.60	<5.00e-02	1.00e+03	1.00e+03	n/a	n/a	97.90	97.90	1.00e+03	n/a	n/a
S981002920	F	Vanadium - ICP-Fusion	ug/g	105.5	<9.30e-03	4.01e-03	4.01e-03	n/a	n/a	n/a	n/a	8.00e-03	n/a	n/a
S981002920	F	Concentration by GEA	ug/g	111.7	5.00e-03	6.14e-03	6.14e-03	n/a	n/a	n/a	n/a	6.14e-03	n/a	n/a
S981002928	A	Sn117 ICP/MS Acid Dig LA-505159	ug/g	98.00	<2.00e-05	2.79e-01	2.79e-01	2.69e-01	4.45	n/a	n/a	8.20e-02	n/a	n/a
S981002928	A	Sn133 ICP/MS Acid Dig LA-505159	ug/g	n/a	4.84e-05	8.61e-02	9.16e-02	8.88e-02	6.19	n/a	n/a	8.20e-02	n/a	n/a
S981002928	A	Sn135 ICP/MS Acid Dig LA-505159	ug/g	n/a	8.80e-05	2.79e-01	2.85e-01	2.81e-01	1.42	n/a	n/a	8.20e-02	n/a	n/a
S981002928	A	Protactinium-231 by ICP/MS	ug/g	n/a	4.02e-05	1.51e-01	1.41e-01	n/a	n/a	n/a	n/a	1.51e-01	n/a	n/a
S981002928	A	Thorium-229 by ICP/MS	ug/g	n/a	<2.52e-05	9.41e-02	4.76e-02	n/a	n/a	n/a	n/a	9.40e-02	n/a	n/a
S981002928	A	Thorium-230 by ICP/MS	ug/g	n/a	<2.17e-05	5.28e-01	4.59e-02	n/a	n/a	n/a	n/a	7.90e-02	n/a	n/a
S981002928	A	Thorium-232 by ICP/MS	ug/g	107.0	<2.77e-05	1.02e-01	4.59e-02	4.94e-01	14.0	n/a	n/a	9.40e-02	n/a	n/a
S981002928	A	Pu241m-241 by ICP/MS	ug/g	104.3	<2.74e-05	4.01e-02	4.01e-02	n/a	n/a	n/a	n/a	1.02e-01	n/a	n/a
S981002928	A	Pu243/244 by ICP/MS	ug/g	n/a	<1.18e-05	4.40e-02	4.01e-02	n/a	n/a	n/a	n/a	4.40e-02	n/a	n/a
S981002928	A	Pu244/244 by ICP/MS	ug/g	n/a	<2.04e-05	4.01e-02	4.01e-02	n/a	n/a	n/a	n/a	7.60e-02	n/a	n/a
S981002928	A	Sliver - ICP-Acid Digest	ug/g	85.00	<1.00e-02	1.82e-01	1.82e-01	18.40	2.17	94.20	94.20	10.20	10.20	n/a
S981002928	A	Arsenic - ICP-Acid Digest	ug/g	90.60	<5.00e-03	1.02e+02	<9.50e-01	n/a	n/a	n/a	n/a	10.20	n/a	n/a
S981002928	A	Beryllium - ICP-Acid Digest	ug/g	109.6	<5.00e-03	<5.100	4.75e-01	n/a	n/a	100.2	100.2	5.100	5.100	n/a
S981002928	A	Cerium - ICP-Acid Digest	ug/g	95.20	<1.00e-01	1.02e+02	<9.50e-01	n/a	n/a	99.80	99.80	10.20	10.20	n/a
S981002928	A	Cobalt - ICP-Acid Digest	ug/g	92.40	<2.00e-02	<20.40	<9.50e-01	n/a	n/a	99.80	99.80	20.40	20.40	n/a
S981002928	A	Copper - ICP-Acid Digest	ug/g	87.00	<1.00e-02	<10.20	<9.50e-01	n/a	n/a	97.60	97.60	10.20	10.20	n/a
S981002928	A	Lanthanum - ICP-Acid Digest	ug/g	94.80	<5.00e-02	<51.00	<9.50e-01	n/a	n/a	97.40	97.40	51.00	51.00	n/a
S981002928	A	Magnesium - ICP-Acid Digest	ug/g	94.00	<1.00e-01	1.02e+02	<9.50e-01	n/a	n/a	102.8	102.8	10.20	10.20	n/a
S981002928	A	Molybdenum - ICP-Acid Digest	ug/g	92.60	<5.00e-02	<51.00	<9.50e-01	n/a	n/a	95.40	95.40	10.20	10.20	n/a
S981002928	A	Nickel - ICP-Acid Digest	ug/g	93.40	<1.00e-01	1.02e+02	<9.50e-01	n/a	n/a	n/a	n/a	10.20	10.20	n/a

Sample#	RI#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002928	A	Phosphorus -ICP-Acid Digest	ug/g	89.40	<2.00e-01	2.98e-02	1.97e-01	352.5	20.8	110.0	204.0	D/A
S981002928	A	Lead -ICP-Acid Digest	ug/g	88.60	<1.00e-01	<1.00e-02	<9.50e-03	N/A	N/A	130.6	102.0	D/A
S981002928	A	Sulfur -ICP-Acid Digest	ug/g	88.00	<1.00e-01	1.09e-02	1.07e-03	1.06e-03	1.89	133.6	102.0	D/A
S981002928	A	Antimony -ICP-Acid Digest	ug/g	88.20	<6.00e-02	9.1e-03	5.70e-03	N/A	N/A	100.6	61.00	D/A
S981002928	A	Selenium -ICP-Acid Digest	ug/g	90.20	<1.00e-01	<1.00e-02	<9.50e-03	N/A	N/A	98.80	102.0	D/A
S981002928	A	Samarium -ICP-Acid Digest	ug/g	92.80	<1.00e-01	<1.00e-02	<9.50e-03	N/A	N/A	95.00	102.0	D/A
S981002928	A	Strontium -ICP-Acid Digest	ug/g	94.80	<1.00e-02	<10.20	<9.50e-03	N/A	N/A	97.00	10.20	D/A
S981002928	A	Thallium -ICP-Acid Digest	ug/g	93.90	<1.00e-02	<10.20	<9.50e-03	N/A	N/A	98.80	10.20	D/A
S981002928	A	Thallium -ICP-Acid Digest	ug/g	93.90	<2.00e-01	<2.04e-02	<1.98e-02	N/A	N/A	93.40	204.0	D/A
S981002928	A	Vanadium -ICP-Acid Digest	ug/g	92.90	<5.00e-02	<5.100	<4.78e-01	N/A	N/A	97.40	51.00	D/A
S981002928	A	Acetate by IC-DIONE	ug/g	91.80	<2.00e-02	59.44	49.10	64.27	15.1	92.74	23.78	D/A
S981002923	W	Glycolate IC-DioneX 4000/4500	ug/g	96.26	<5.00e-02	<65.57	<6.68e-01	N/A	N/A	93.18	66.60	D/A

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

25-mar-1999 10:55:20
A-0002-1

CORE NUMBER: 250
SEGMENT #: 7

SEGMENT PORTION: L Lower Half of Segment

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicates	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002921	F	Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.01e-02	<2.03e2	N/A	N/A	N/A	201.0	N/A
S98T002921	F	Arsenic - ICP-Fusion	ug/g	101.0	<1.00e-01	<2.01e-03	<2.03e3	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Beryllium - ICP-Fusion	ug/g	101.8	<5.00e-03	<1.01e-02	<1.01e2	N/A	N/A	N/A	100.0	N/A
S98T002921	F	Cadmium - ICP-Fusion	ug/g	103.0	<5.00e-03	<1.01e-02	<2.03e2	N/A	N/A	N/A	100.0	N/A
S98T002921	F	Cerium - ICP-Fusion	ug/g	98.80	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Cobalt - ICP-Fusion	ug/g	100.0	<2.00e-02	<4.02e-03	<2.03e2	N/A	N/A	N/A	402.0	N/A
S98T002921	F	Copper - ICP-Fusion	ug/g	92.40	<1.00e-02	<2.01e-03	<2.03e2	N/A	N/A	N/A	201.0	N/A
S98T002921	F	Lead - ICP-Fusion	ug/g	99.00	<5.00e-02	<1.01e-03	<2.03e2	N/A	N/A	N/A	1.00e+03	N/A
S98T002921	F	Lanthanum - ICP-Fusion	ug/g	102.4	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Magnesium - ICP-Fusion	ug/g	101.2	<5.00e-02	<1.01e-03	<2.03e2	N/A	N/A	N/A	1.00e+03	N/A
S98T002921	F	Molybdenum - ICP-Fusion	ug/g	96.00	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Neodymium - ICP-Fusion	ug/g	101.0	<2.00e-01	<4.02e-03	<2.03e2	N/A	N/A	N/A	4.02e+03	N/A
S98T002921	F	Phosphorus - ICP-Fusion	ug/g	101.0	<2.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Silver - ICP-Fusion	ug/g	99.80	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Sulfur - ICP-Fusion	ug/g	101.2	<1.00e-02	<1.01e-03	<2.03e2	N/A	N/A	N/A	1.00e+03	N/A
S98T002921	F	Tantalum - ICP-Fusion	ug/g	97.00	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Selenium - ICP-Fusion	ug/g	97.00	<1.00e-01	<2.01e-03	<2.03e2	N/A	N/A	N/A	2.01e+03	N/A
S98T002921	F	Strontium - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.01e-03	<2.03e2	N/A	N/A	N/A	201.0	N/A
S98T002921	F	Titanium - ICP-Fusion	ug/g	96.00	<2.00e-01	<4.02e-03	<2.03e2	N/A	N/A	N/A	4.02e+03	N/A
S98T002921	F	Thallium - ICP-Fusion	ug/g	99.60	<5.00e-02	<1.01e-03	<1.01e3	N/A	N/A	N/A	1.00e+03	N/A
S98T002921	F	Vanadium - ICP-Fusion	ug/g	105.5	<9.30e-03	<8.43e-03	<8.05e-3	N/A	N/A	N/A	8.00e-03	N/A
S98T002921	F	Cobalt-60 by GEA	ug/g	111.7	5.00e-03	< 5.930	<6.13e0	N/A	N/A	N/A	5.930	N/A
S98T002929	A	Sn117 ICP/MS Actidig LA-505159	ug/g	98.00	<2.00e-05	1.218	1.230	1.224	0.82	N/A	7.90e-02	N/A
S98T002929	A	Cs132 ICP/MS Actidig LA-505159	ug/g	N/A	4.84e-05	4.70e-01	4.66e-01	4.68e-01	0.85	N/A	7.90e-02	N/A
S98T002929	A	Cs137 ICP/MS Actidig LA-505159	ug/g	N/A	8.80e-05	6.29e-01	6.21e-01	6.25e-01	1.28	N/A	7.90e-02	N/A
S98T002929	A	Cs137 ICP/MS Actidig LA-505159	ug/g	N/A	4.84e-05	<1.46e-01	<1.51e-1	<1.51e-1	N/A	N/A	1.46e-01	N/A
S98T002929	A	Protactinium-231 by ICP/MS	ug/g	N/A	<2.52e-05	<9.07e-02	<9.38e-2	N/A	N/A	N/A	9.10e-02	N/A
S98T002929	A	Thorium-232 by ICP/MS	ug/g	N/A	<2.13e-05	<9.07e-02	<9.92e-2	N/A	N/A	N/A	7.10e-02	N/A
S98T002929	A	Thorium-232 by ICP/MS	ug/g	107.0	<2.77e-05	<6.30e-01	6.95e-01	6.63e-01	9.81	N/A	9.00e-02	N/A
S98T002929	A	Thorium-232 by ICP/MS	ug/g	104.3	<2.77e-05	<9.87e-02	<1.02e-1	N/A	N/A	N/A	9.90e-02	N/A
S98T002929	A	U-235/238 by ICP/MS	ug/g	N/A	<1.18e-05	<4.24e-02	<4.38e-2	N/A	N/A	N/A	4.20e-02	N/A
S98T002929	A	U-235/238 by ICP/MS	ug/g	N/A	<2.04e-05	<9.32e-02	<9.56e-2	N/A	N/A	N/A	7.30e-02	N/A
S98T002929	A	Silver - ICP-Acid Digest	ug/g	85.00	<1.00e-02	<1.02e2	19.50	17.40	18.45	11.4	9.850	N/A
S98T002929	A	Beryllium - ICP-Acid Digest	ug/g	90.60	<5.00e-03	< 4.920	<5.08e0	N/A	N/A	N/A	98.50	N/A
S98T002929	A	Cerium - ICP-Acid Digest	ug/g	100.6	<5.00e-03	< 4.920	<5.08e0	N/A	N/A	N/A	4.920	N/A
S98T002929	A	Cobalt - ICP-Acid Digest	ug/g	91.80	<5.00e-03	< 4.920	<5.08e0	N/A	N/A	N/A	98.50	N/A
S98T002929	A	Copper - ICP-Acid Digest	ug/g	95.20	<2.00e-02	< 9.830	<1.02e1	N/A	N/A	N/A	19.70	N/A
S98T002929	A	Lanthanum - ICP-Acid Digest	ug/g	92.40	<2.00e-02	< 9.830	<1.02e1	N/A	N/A	N/A	9.850	N/A
S98T002929	A	Magnesium - ICP-Acid Digest	ug/g	87.00	<1.00e-02	< 4.920	<5.08e1	N/A	N/A	N/A	49.10	N/A
S98T002929	A	Molybdenum - ICP-Acid Digest	ug/g	94.00	<5.00e-02	< 4.920	<5.08e1	N/A	N/A	N/A	98.50	N/A
S98T002929	A	Neodymium - ICP-Acid Digest	ug/g	92.60	<5.00e-02	< 4.920	<5.08e1	N/A	N/A	N/A	49.10	N/A
S98T002929	A	Neodymium - ICP-Acid Digest	ug/g	93.40	<1.00e-01	< 98.30	<1.02e2	N/A	N/A	N/A	98.50	N/A

5090

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S981002929	A	Phosphorus -ICP-Acid Digest	ug/g	88.40	<2.00e-01	4.02e-02	155.3	427.0	11.7	194.0	N/A
S981002929	A	Lead -ICP-Acid Digest	ug/g	88.60	<1.00e-01	1.98e-01	105.3	985.0	N/A	98.50	N/A
S981002929	A	Sulfur -ICP-Acid Digest	ug/g	88.00	<1.00e-01	1.03e-01	942.6	985.0	8.92	98.50	N/A
S981002929	A	Antimony -ICP-Acid Digest	ug/g	88.20	<6.00e-02	1.55e-01	45.16	N/A	N/A	59.00	N/A
S981002929	A	Selenium -ICP-Acid Digest	ug/g	90.20	<1.00e-01	98.30	1.02e2	N/A	N/A	98.50	N/A
S981002929	A	Samarium -ICP-Acid Digest	ug/g	92.60	<1.00e-01	98.30	1.02e2	N/A	N/A	98.50	N/A
S981002929	A	Strontium -ICP-Acid Digest	ug/g	94.80	<1.00e-02	9.830	1.02e1	N/A	N/A	9.850	N/A
S981002929	A	Thallium -ICP-Acid Digest	ug/g	93.80	<1.00e-02	1.97e-02	2.03e2	N/A	N/A	9.850	N/A
S981002929	A	Vanadium -ICP-Acid Digest	ug/g	92.30	<5.00e-02	4.70e-02	5.08e1	N/A	N/A	49.10	N/A
S981002934	W	Acetate by IC-DIONEX 4000/4500	ug/g	100.0	2.80e-02	4.20e-02	382.0	401.1	9.48	23.45	N/A
S981002934	W	Glycolate IC-DIONEX 4000/4500	ug/g	96.26	<5.60e-02	65.25	-65.61e1	N/A	N/A	65.64	N/A

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 250
SEGMENT #: 98

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk	Rec %	Det Limit	Count	Err%
S981002922	F	Silver - ICP-Fusion	ug/g	99.80	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	87.50	209.0	n/a	n/a
S981002922	F	Arsenic - ICP-Fusion	ug/g	101.0	<1.00e-01	<4.06e-03	<4.06e-03	n/a	n/a	n/a	97.40	2.09e+03	n/a	n/a
S981002922	F	Beryllium - ICP-Fusion	ug/g	101.8	<5.00e-03	<1.09e-02	<1.09e-02	n/a	n/a	n/a	94.50	105.0	n/a	n/a
S981002922	F	Cadmium - ICP-Fusion	ug/g	98.60	<5.00e-03	<2.09e-02	<2.09e-02	n/a	n/a	n/a	100.0	105.0	n/a	n/a
S981002922	F	Cerium - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	90.40	2.09e+03	n/a	n/a
S981002922	F	Cobalt - ICP-Fusion	ug/g	98.60	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	97.10	4.18e-01	n/a	n/a
S981002922	F	Copper - ICP-Fusion	ug/g	98.60	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	84.80	209.0	n/a	n/a
S981002922	F	Lanthanum - ICP-Fusion	ug/g	98.60	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	90.70	1.05e+03	n/a	n/a
S981002922	F	Magnesium - ICP-Fusion	ug/g	102.4	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	98.70	2.09e+03	n/a	n/a
S981002922	F	Molybdenum - ICP-Fusion	ug/g	101.2	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	87.70	1.05e+03	n/a	n/a
S981002922	F	Neodymium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	87.70	2.09e+03	n/a	n/a
S981002922	F	Phosphorus - ICP-Fusion	ug/g	101.0	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	87.70	1.05e+03	n/a	n/a
S981002922	F	Lead - ICP-Fusion	ug/g	98.60	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	101.0	2.09e+03	n/a	n/a
S981002922	F	Sulfur - ICP-Fusion	ug/g	98.60	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	87.70	1.05e+03	n/a	n/a
S981002922	F	Antimony - ICP-Fusion	ug/g	97.00	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	97.70	2.09e+03	n/a	n/a
S981002922	F	Selenium - ICP-Fusion	ug/g	97.00	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	97.70	2.09e+03	n/a	n/a
S981002922	F	Samarium - ICP-Fusion	ug/g	97.00	<1.00e-01	<2.09e-02	<2.09e-02	n/a	n/a	n/a	97.70	2.09e+03	n/a	n/a
S981002922	F	Strontium - ICP-Fusion	ug/g	100.8	<1.00e-02	<2.09e-02	<2.09e-02	n/a	n/a	n/a	90.10	2.09e+03	n/a	n/a
S981002922	F	Thallium - ICP-Fusion	ug/g	96.60	<2.00e-01	<4.18e-03	<4.18e-03	n/a	n/a	n/a	95.90	1.05e+03	n/a	n/a
S981002922	F	Vanadium - ICP-Fusion	ug/g	96.60	<5.00e-03	<8.48e-02	<8.48e-02	n/a	n/a	n/a	88.30	209.0	n/a	n/a
S981002922	F	Cobalt-60 by GEA	ug/g	111.7	5.00e-03	<2.44e-02	<2.44e-02	n/a	n/a	n/a	n/a	240.3	n/a	n/a
S981002930	A	SH17 ICP/MS Act100/g LA-508159	ug/g	98.00	<2.00e-05	<3.70e-03	<3.70e-03	n/a	n/a	n/a	n/a	3.204	n/a	n/a
S981002930	A	GS13 ICP/MS Act100/g LA-508159	ug/g	n/a	4.84e-05	<3.20e-03	<3.20e-03	n/a	n/a	n/a	n/a	3.204	n/a	n/a
S981002930	A	GS13 ICP/MS Act100/g LA-508159	ug/g	n/a	4.84e-05	<3.20e-03	<3.20e-03	n/a	n/a	n/a	n/a	3.204	n/a	n/a
S981002930	A	Protactinium-231 by CP/MS	ug/g	n/a	4.84e-05	<3.20e-03	<3.20e-03	n/a	n/a	n/a	n/a	5.912	n/a	n/a
S981002930	A	Thorium-230 by CP/MS	ug/g	n/a	4.84e-05	<3.20e-03	<3.20e-03	n/a	n/a	n/a	n/a	3.672	n/a	n/a
S981002930	A	Thorium-232 by CP/MS	ug/g	n/a	4.84e-05	<3.20e-03	<3.20e-03	n/a	n/a	n/a	n/a	3.104	n/a	n/a
S981002930	A	Thorium-232 by CP/MS	ug/g	107.0	<2.77e-05	<3.651	<3.651	n/a	n/a	n/a	98.50	3.661	n/a	n/a
S981002930	A	Thorium-232 by CP/MS	ug/g	104.3	<2.77e-05	<3.651	<3.651	n/a	n/a	n/a	98.50	3.661	n/a	n/a
S981002930	A	Plutonium-241 by CP/MS	ug/g	n/a	<1.18e-05	<1.719	<1.719	n/a	n/a	n/a	n/a	1.719	n/a	n/a
S981002930	A	Plutonium-241 by CP/MS	ug/g	n/a	<1.18e-05	<1.719	<1.719	n/a	n/a	n/a	n/a	2.966	n/a	n/a
S981002930	A	Plutonium-241 by CP/MS	ug/g	85.00	<1.00e-02	<1.920	<1.920	n/a	n/a	n/a	94.10	97.00	n/a	n/a
S981002930	A	Silver - ICP-Acid Digest	ug/g	90.60	<1.00e-01	<97.20	<97.20	n/a	n/a	n/a	101.2	97.00	n/a	n/a
S981002930	A	Arsenic - ICP-Acid Digest	ug/g	100.6	<5.00e-03	<4.860	<4.860	n/a	n/a	n/a	102.2	4.860	n/a	n/a
S981002930	A	Beryllium - ICP-Acid Digest	ug/g	95.20	<1.00e-01	<97.20	<97.20	n/a	n/a	n/a	100.4	97.00	n/a	n/a
S981002930	A	Cadmium - ICP-Acid Digest	ug/g	95.20	<1.00e-01	<97.20	<97.20	n/a	n/a	n/a	100.4	97.00	n/a	n/a
S981002930	A	Cerium - ICP-Acid Digest	ug/g	92.40	<1.00e-02	<9.720	<9.720	n/a	n/a	n/a	85.00	97.00	n/a	n/a
S981002930	A	Cobalt - ICP-Acid Digest	ug/g	87.00	<1.00e-02	<4.860	<4.860	n/a	n/a	n/a	99.20	48.60	n/a	n/a
S981002930	A	Copper - ICP-Acid Digest	ug/g	94.00	<5.00e-02	<7.09e+02	<7.09e+02	n/a	n/a	n/a	81.80	97.00	n/a	n/a
S981002930	A	Lanthanum - ICP-Acid Digest	ug/g	94.00	<1.00e-01	<48.60	<48.60	n/a	n/a	n/a	101.0	48.60	n/a	n/a
S981002930	A	Magnesium - ICP-Acid Digest	ug/g	92.60	<5.00e-02	<48.60	<48.60	n/a	n/a	n/a	100.0	97.00	n/a	n/a
S981002930	A	Molybdenum - ICP-Acid Digest	ug/g	93.40	<1.00e-01	<97.20	<97.20	n/a	n/a	n/a	100.0	97.00	n/a	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002930	A	Phosphorus -ICP-Acid Digest	ug/g	89.40	< 2.00e-01	9.90e+03	7.27e+03	8.58e+03	30.6	-1.04e-02	194.0	n/a
S981002930	A	Lead -ICP-Acid Digest	ug/g	88.00	< 1.00e-01	2.38e+02	222.0	255.0	25.9	93.00	97.00	n/a
S981002930	A	Sulfur -ICP-Acid Digest	ug/g	86.00	< 1.00e-01	8.36e+03	2.71e+03	5.54e+03	102	-5.22e-02	97.00	n/a
S981002930	A	Antimony -ICP-Acid Digest	ug/g	88.20	< 6.00e-02	< 58.30	< 5.961	n/a	n/a	100.2	58.50	n/a
S981002930	A	Selenium -ICP-Acid Digest	ug/g	90.20	< 1.00e-01	< 97.20	< 9.881	n/a	n/a	100.2	97.00	n/a
S981002930	A	Samarium -ICP-Acid Digest	ug/g	92.80	< 1.00e-01	< 97.20	< 9.881	n/a	n/a	96.80	97.00	n/a
S981002930	A	Strontium -ICP-Acid Digest	ug/g	94.00	< 1.00e-02	3.92e+03	3.36e+03	3.64e+03	15.6	46.00	9.700	n/a
S981002930	A	Titanium -ICP-Acid Digest	ug/g	92.80	< 1.00e-02	57.20	54.60	55.90	4.65	99.20	9.700	n/a
S981002930	A	Thallium -ICP-Acid Digest	ug/g	87.80	< 2.00e-01	< 1.94e+02	< 1.982	n/a	n/a	93.60	194.0	n/a
S981002935	W	Vanadium -ICP-Acid Digest	ug/g	92.20	< 5.00e-02	48.60	44.941	n/a	n/a	99.40	48.60	n/a
S981002935	W	Acetate by IC-DIONEX 4000/4500	ug/g	100.0	< 2.80e-02	4.65e+02	569.0	517.0	20.1	96.55	24.19	n/a
S981002935	W	Glycolate IC-Dionex 4000/4500	ug/g	96.26	< 5.60e-02	< 67.74	< 6.76e1	n/a	n/a	94.39	67.74	n/a

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002939	D	Silver-ICP-Acid Dill.	ug/ml	99.60	< 1.00e-02	14.10	14.20	14.15	0.71	79.70	6.010	n/a
S981002939	D	Arsenic-ICP-Acid Dill.	ug/ml	101.4	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	97.10	60.10	n/a
S981002939	D	Beryllium-ICP-Acid Dill.	ug/ml	101.4	< 5.00e-01	< 3.010	< 3.010	n/a	n/a	98.00	3.000	n/a
S981002939	D	Cadmium-ICP-Acid Dill.	ug/ml	102.0	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	99.20	3.000	n/a
S981002939	D	Cerium-ICP-Acid Dill.	ug/ml	99.80	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	94.50	60.10	n/a
S981002939	D	Cobalt-ICP-Acid Dill.	ug/ml	99.20	< 2.00e-02	< 12.00	< 1.20e1	n/a	n/a	96.40	12.00	n/a
S981002939	D	Copper-ICP-Acid Dill.	ug/ml	94.00	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	89.50	6.010	n/a
S981002939	D	Lanthanum-ICP-Acid Dill.	ug/ml	100.2	< 5.00e-02	< 30.10	< 3.010	n/a	n/a	96.70	30.10	n/a
S981002939	D	Magnesium-ICP-Acid Dill.	ug/ml	100.4	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	96.50	60.10	n/a
S981002939	D	Molybdenum-ICP-Acid Dill.	ug/ml	100.4	< 5.00e-02	< 35.10	< 3.510	35.30	1.13	97.60	30.10	n/a
S981002939	D	Neodymium-ICP-Acid Dill.	ug/ml	97.60	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	93.40	60.10	n/a
S981002939	D	Phosphorus-ICP-Acid Dill.	ug/ml	101.2	< 2.00e-01	5.44e+02	442.0	503.0	16.3	81.30	120.0	n/a
S981002939	D	Sulfur-ICP-Acid Dill.	ug/ml	98.60	< 1.00e-01	87.70	89.50	88.60	2.03	99.60	60.10	n/a
S981002939	D	Sulfur-ICP-Acid Dill.	ug/ml	98.60	< 6.00e-02	7.77e+02	779.0	778.0	0.26	97.10	60.10	n/a
S981002939	D	Selenium-ICP-Acid Dill.	ug/ml	101.8	< 6.00e-02	< 35.10	< 3.510	n/a	n/a	100.0	35.10	n/a
S981002939	D	Samarium-ICP-Acid Dill.	ug/ml	96.60	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	105.0	60.10	n/a
S981002939	D	Samarium-ICP-Acid Dill.	ug/ml	97.40	< 1.00e-01	< 60.10	< 6.010	n/a	n/a	96.00	60.10	n/a
S981002939	D	Strontium-ICP-Acid Dill.	ug/ml	97.80	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	83.90	6.010	n/a
S981002939	D	Titanium-ICP-Acid Dill.	ug/ml	101.2	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	97.00	6.010	n/a
S981002939	D	Thallium-ICP-Acid Dill.	ug/ml	96.40	< 2.00e-02	< 30.10	< 3.010	n/a	n/a	91.80	30.10	n/a
S981002939	D	Vanadium-ICP-Acid Dill.	ug/ml	99.20	< 5.00e-02	< 30.10	< 3.010	n/a	n/a	96.00	30.10	n/a
S981002939	D	Acetate by IC-Dionex 4000/4500	ug/ml	100.9	< 2.00e-02	2.37e+03	2.37e+03	2.37e+03	0.00	109.4	36.10	n/a
S981002939	D	Glycolate-IC-Dionex 4500 ORGAC	ug/ml	100.9	< 2.00e-02	< 33.94	< 3.394	n/a	n/a	107.4	33.94	n/a
S981002939	D	Cobalt-60 by BEA	uCi/ml	100.3	< 1.24e-03	< 4.02e-03	< 4.02e-03	3.08e+03	2.68	n/a	4.00e-03	n/a
S981002940	D	TiC by Acid/Coulometry	ug/ml	100.8	< 7.00e-01	3.11e+03	3.06e+03	3.08e+03	104.0	n/a	3.000	n/a
S981002940	D	Sn 117 by ICP/MS Acid Addition	ug/ml	81.82	< 1.50e-04	< 0.075	< 0.080	n/a	n/a	n/a	6.073	n/a
S981002940	D	Cs 133 by ICP/MS Acid Addition	ug/ml	105.5	< 1.24e-05	7.023	9.880	6.926	2.74	105.0	8.10e-02	n/a
S981002940	D	Cs 135 by ICP/MS Acid Addition	ug/ml	n/a	< 2.00e-06	2.210	2.110	2.160	4.63	n/a	8.10e-02	n/a
S981002940	D	Cs 137 by ICP/MS Acid Addition	ug/ml	n/a	< 2.00e-06	1.847	1.720	1.798	5.56	n/a	8.10e-02	n/a
S981002940	D	Protactinium-231 by ICP/MS	ug/ml	n/a	< 1.54e-05	< 6.25e-01	< 6.25e-01	n/a	n/a	n/a	6.25e-01	n/a
S981002940	A	Thorium-229 by ICP/MS	ug/ml	n/a	n/a	1.268	< 1.27e0	n/a	n/a	n/a	1.268	n/a
S981002940	A	Thorium-230 by ICP/MS	ug/ml	96.10	< 8.70e-01	< 8.69e-01	< 8.69e-01	n/a	n/a	n/a	8.70e-01	n/a
S981002940	A	Thorium-232 by ICP/MS	ug/ml	n/a	< 1.58e-02	< 6.41e-01	< 6.41e-01	n/a	n/a	86.67	7.86e-01	n/a
S981002940	A	Pu/Am-241 by ICP/MS	ug/ml	0.23	< 1.58e-02	< 6.41e-01	< 6.41e-01	n/a	n/a	2.54e-01	6.41e-01	n/a
S981002940	A	Am 243 (m 243 by ICP/MS	ug/ml	n/a	< 1.44e-05	< 5.85e-01	< 5.85e-01	n/a	n/a	n/a	5.85e-01	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	% Spk Rec	Det Limit	Count Err%
S98T02240	A	Pu 244/Cm 244 by ICP/MS	ug/mL	N/A	<1.22e-05	<4.96e-01	<4.96e-1	N/A	N/A	N/A	4.96e-01	N/A

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 250

SEGMENT #: Core Composite

SEGMENT PORTION: Core Composite

Sample #	RI#	Analyte	Unit	Standard %	Blank	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count	Err%
S98T003388	F	Silver - ICP-Fusion	ug/g	98.00	< 1.00e-02	< 2.03e-02	N/A	N/A	90.30	203.0	N/A	N/A
S98T003388	F	Arsenic - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.03e-03	N/A	N/A	101.0	2.03e+03	N/A	N/A
S98T003388	F	Beryllium - ICP-Fusion	ug/g	101.0	< 5.00e-03	< 1.01e-02	N/A	N/A	96.40	101.0	N/A	N/A
S98T003388	F	Barium - ICP-Fusion	ug/g	100.0	< 1.00e-03	< 1.01e-02	N/A	N/A	101.0	101.0	N/A	N/A
S98T003388	F	Cerium - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.03e-03	N/A	N/A	96.30	2.03e+03	N/A	N/A
S98T003388	F	Cobalt - ICP-Fusion	ug/g	100.0	< 2.00e-02	< 4.05e-02	N/A	N/A	99.30	203.0	N/A	N/A
S98T003388	F	Copper - ICP-Fusion	ug/g	95.00	< 1.00e-02	< 2.03e-03	N/A	N/A	90.10	203.0	N/A	N/A
S98T003388	F	Lead - ICP-Fusion	ug/g	100.0	< 5.00e-02	< 1.01e-03	N/A	N/A	94.90	1.01e+03	N/A	N/A
S98T003388	F	Lanthanum - ICP-Fusion	ug/g	103.0	< 1.00e-01	< 2.03e-03	N/A	N/A	99.40	2.03e+03	N/A	N/A
S98T003388	F	Magnesium - ICP-Fusion	ug/g	105.0	< 5.00e-02	< 1.01e-03	N/A	N/A	104.0	1.01e+03	N/A	N/A
S98T003388	F	Molybdenum - ICP-Fusion	ug/g	97.40	< 1.00e-01	< 2.03e-03	N/A	N/A	91.70	2.03e+03	N/A	N/A
S98T003388	F	Nickel - ICP-Fusion	ug/g	105.0	< 2.00e-01	< 4.05e-03	N/A	N/A	114.0	4.05e+03	N/A	N/A
S98T003388	F	Phosphorus - ICP-Fusion	ug/g	100.0	< 1.00e-01	< 2.03e-03	N/A	N/A	102.0	2.03e+03	N/A	N/A
S98T003388	F	Silver - ICP-Fusion	ug/g	102.0	< 1.00e-01	< 2.03e-03	N/A	N/A	111.0	1.22e+03	N/A	N/A
S98T003388	F	Selenium - ICP-Fusion	ug/g	106.0	< 5.00e-02	< 1.01e-03	N/A	N/A	106.0	1.22e+03	N/A	N/A
S98T003388	F	Selenium - ICP-Fusion	ug/g	96.40	< 1.00e-01	< 2.03e-03	N/A	N/A	98.90	2.03e+03	N/A	N/A
S98T003388	F	Selenium - ICP-Fusion	ug/g	97.40	< 1.00e-01	< 2.03e-03	N/A	N/A	92.40	2.03e+03	N/A	N/A
S98T003388	F	Strontium - ICP-Fusion	ug/g	106.0	< 1.00e-02	< 2.03e-02	N/A	N/A	102.0	203.0	N/A	N/A
S98T003388	F	Thallium - ICP-Fusion	ug/g	95.20	< 2.00e-02	< 4.05e-02	N/A	N/A	92.00	4.05e+03	N/A	N/A
S98T003388	F	Vanadium - ICP-Fusion	ug/g	100.0	< 5.00e-02	< 1.01e-03	N/A	N/A	98.40	1.01e+03	N/A	N/A
S98T003388	F	Cobalt-60 by GEA	uS/g	115.3	< 4.00e-03	< 6.18e-03	N/A	N/A	108.0	9.00e+03	N/A	N/A
S98T003389	A	Ca117 ICP/MS Acid Dig LA-505159	ug/g	99.00	< 2.00e-03	5.27e-01	5.23e-01	1.53	101.5	8.20e-02	N/A	N/A
S98T003389	A	Ca133 ICP/MS Acid Dig LA-505159	ug/g	N/A	4.30e-05	2.47e-01	2.32e-01	12.9	N/A	8.20e-02	N/A	N/A
S98T003389	A	Ca137 ICP/MS Acid Dig LA-505159	ug/g	N/A	6.80e-05	5.69e-01	5.66e-01	0.88	N/A	8.20e-02	N/A	N/A
S98T003389	A	Protectinilm-231 by ICP/MS	ug/g	N/A	5.57e-05	< 3.0e-01	< 2.12e-1	N/A	N/A	2.30e-01	N/A	N/A
S98T003389	A	Thorium-229 by ICP/MS	ug/g	N/A	4.44e-05	< 1.8e-01	< 1.70e-1	N/A	N/A	1.84e-01	N/A	N/A
S98T003389	A	Thorium-230 by ICP/MS	ug/g	N/A	3.52e-05	< 1.4e-01	< 1.39e-1	N/A	N/A	1.46e-01	N/A	N/A
S98T003389	A	Thorium-232 by ICP/MS	ug/g	103.6	2.12e-05	1.954	2.027	7.41	97.50	8.00e-02	N/A	N/A
S98T003389	A	Pu/Am-241 by ICP/MS	ug/g	107.8	3.71e-05	< 1.53e-01	< 1.41e-1	N/A	104.3	1.53e-01	N/A	N/A
S98T003389	A	Am-243/Am-243 by ICP/MS	ug/g	N/A	3.83e-05	< 1.58e-01	< 1.46e-1	N/A	N/A	1.58e-01	N/A	N/A
S98T003389	A	Pu-244/Am-244 by ICP/MS	ug/g	N/A	3.41e-05	< 1.41e-01	< 1.30e-1	N/A	N/A	1.41e-01	N/A	N/A
S98T003389	A	Silver - ICP-Acid Digest	ug/g	85.00	< 1.00e-02	< 9.48e-1	17.35	9.80	94.70	10.20	N/A	N/A
S98T003389	A	Arsenic - ICP-Acid Digest	ug/g	92.20	< 5.00e-03	< 1.03e+02	< 9.48e-1	N/A	99.20	102.0	N/A	N/A
S98T003389	A	Beryllium - ICP-Acid Digest	ug/g	103.0	< 5.00e-03	< 5.130	< 4.74e0	N/A	100.6	5.150	N/A	N/A
S98T003389	A	Barium - ICP-Acid Digest	ug/g	93.00	< 5.00e-03	< 5.130	< 4.74e0	N/A	99.20	5.150	N/A	N/A
S98T003389	A	Cerium - ICP-Acid Digest	ug/g	93.40	< 1.00e-01	< 1.03e+02	< 9.48e-1	N/A	97.00	102.0	N/A	N/A
S98T003389	A	Cobalt - ICP-Acid Digest	ug/g	97.00	< 2.00e-02	< 20.50	< 1.90e-1	N/A	99.00	20.50	N/A	N/A
S98T003389	A	Copper - ICP-Acid Digest	ug/g	88.40	< 1.00e-02	< 10.30	< 9.48e0	N/A	89.80	10.20	N/A	N/A
S98T003389	A	Lanthanum - ICP-Acid Digest	ug/g	97.00	< 5.00e-02	< 51.30	< 4.74e0	N/A	96.60	51.50	N/A	N/A
S98T003389	A	Magnesium - ICP-Acid Digest	ug/g	94.00	< 1.00e-01	< 1.03e+02	< 9.48e-1	N/A	102.6	102.0	N/A	N/A
S98T003389	A	Molybdenum - ICP-Acid Digest	ug/g	95.00	< 5.00e-02	< 51.30	< 4.74e0	N/A	101.0	51.50	N/A	N/A
S98T003389	A	Nickel - ICP-Acid Digest	ug/g	95.80	< 1.00e-01	< 1.03e+02	< 9.48e-1	N/A	94.80	102.0	N/A	N/A

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk	Rec %	Det Limit	Count Err%
S981003389	A	Phosphorus -ICP-Acid Digest	ug/g	81.20	<2.00e-01	1.83e+03	1.67e+03	1.65e+03	21.8	84.90	205.0	N/A
S981003389	A	Lead -ICP-Acid Digest	ug/g	88.80	<1.00e-01	<1.09e+02	1.47e+02	N/A	N/A	102.2	102.0	N/A
S981003389	A	Sulfur -ICP-Acid Digest	ug/g	88.20	1.31e-01	1.74e+03	1.74e+03	1.76e+03	2.83	113.2	102.0	N/A
S981003389	A	Antimony -ICP-Acid Digest	ug/g	91.40	<6.00e-02	<5.69e-01	<5.69e-01	N/A	N/A	100.0	61.50	N/A
S981003389	A	Selenium -ICP-Acid Digest	ug/g	91.20	<1.00e-01	<1.09e+02	<9.48e-01	N/A	N/A	100.4	102.0	N/A
S981003389	A	Samarium -ICP-Acid Digest	ug/g	95.20	<1.00e-01	<1.09e+02	<9.48e-01	N/A	N/A	94.40	102.0	N/A
S981003389	A	Strontium -ICP-Acid Digest	ug/g	96.00	<1.00e-02	<10.50	<9.48e-01	10.20	0.00	96.20	10.20	N/A
S981003389	A	Titanium -ICP-Acid Digest	ug/g	96.40	<1.00e-02	<10.50	<9.48e-01	N/A	N/A	98.80	10.20	N/A
S981003389	A	Thallium -ICP-Acid Digest	ug/g	97.60	<1.00e-01	<2.05e+02	<1.99e+02	N/A	N/A	90.20	205.0	N/A
S981003389	A	Vanadium -ICP-Acid Digest	ug/g	97.40	<1.00e-02	<5.13e-01	<4.74e-01	N/A	N/A	97.00	51.50	N/A
S981003390	W	Acetate by IC-DIONEX 4000/4500	ug/g	92.74	<2.00e-02	1.35e+02	1.62e+02	138.6	5.05	91.13	43.48	N/A
S981003390	W	Glycolate IC-Dionex 4000/4500	ug/g	98.13	<5.60e-02	<1.26e+02	<1.26e+02	N/A	N/A	95.33	122.3	N/A

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 251
SEGMENT #: 1

SEGMENT PORTION: Drainable Liquid

Sample #	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002949	D	Silver-CP-Acid DIL	ug/mL	99.20	<1.00e-02	15.20	14.90	15.05	1.99	91.91	6.010	N/A
S981002949	D	Silver-CP-Acid DIL	ug/mL	100.4	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	97.70	60.10	N/A
S981002949	D	Argent-CP-Acid DIL	ug/mL	102.0	<5.00e-03	3.010	<3.01e0	N/A	N/A	99.91	3.000	N/A
S981002949	D	Beryllium-CP-Acid DIL	ug/mL	102.2	<5.00e-03	3.010	<3.01e0	N/A	N/A	101.0	3.000	N/A
S981002949	D	Cadmium-CP-Acid DIL	ug/mL	99.80	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	97.30	60.10	N/A
S981002949	D	Cerium-CP-Acid DIL	ug/mL	99.20	<2.00e-02	12.00	<1.20e1	N/A	N/A	97.30	12.00	N/A
S981002949	D	Cobalt-CP-Acid DIL	ug/mL	94.00	<1.00e-02	< 60.10	<6.01e0	N/A	N/A	90.40	6.010	N/A
S981002949	D	Copper-CP-Acid DIL	ug/mL	100.4	<5.00e-02	30.10	<3.01e1	N/A	N/A	97.70	30.10	N/A
S981002949	D	Lanthanum-CP-Acid DIL	ug/mL	103.4	<5.00e-01	< 60.10	<6.01e1	N/A	N/A	99.50	60.10	N/A
S981002949	D	Nickel-CP-Acid DIL	ug/mL	100.6	<5.00e-02	31.00	<3.01e1	N/A	N/A	94.50	60.10	N/A
S981002949	D	Polystyrene-CP-Acid DIL	ug/mL	102.4	<2.00e-01	3.42e+02	347.0	344.5	1.45	100.0	120.0	N/A
S981002949	D	Phosphorus-CP-Acid DIL	ug/mL	101.0	<1.00e-01	61.10	72.60	66.85	17.2	101.0	60.10	N/A
S981002949	D	Lead-CP-Acid DIL	ug/mL	97.40	<1.00e-01	7.47e+02	747.0	747.0	0.00	93.10	60.10	N/A
S981002949	D	Antimony-CP-Acid DIL	ug/mL	101.8	<6.00e-02	36.10	<3.61e1	N/A	N/A	102.0	36.10	N/A
S981002949	D	Selenium-CP-Acid DIL	ug/mL	97.00	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	113.0	60.10	N/A
S981002949	D	Strontium-CP-Acid DIL	ug/mL	97.80	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	94.70	60.10	N/A
S981002949	D	Titanium-CP-Acid DIL	ug/mL	98.00	<1.00e-02	< 60.10	<6.01e0	N/A	N/A	94.90	6.010	N/A
S981002949	D	Thallium-CP-Acid DIL	ug/mL	95.60	<2.00e-01	<1.20e+02	<1.20e2	N/A	N/A	91.90	120.0	N/A
S981002949	D	Vanadium-CP-Acid DIL	ug/mL	90.52	<5.00e-02	30.10	<3.01e1	N/A	N/A	97.00	30.10	N/A
S981002949	D	Vanadium-CP-Acid DIL	ug/mL	90.52	<2.00e-02	1.69e+03	1.47e+03	1.58e+03	13.9	68.19	35.15	N/A
S981002949	D	Acetate by IC-Dionex 4000/4500	ug/mL	92.59	<2.00e-02	33.94	<3.39e1	N/A	N/A	105.6	33.94	N/A
S981002949	D	Glycolate-IC-Dionex 4500 ORGAC	ug/mL	110.3	<8.53e-04	7.26e+03	7.05e+03	7.16e+03	2.94	112.0	5.000	N/A
S981002949	D	Cobalt-60 by GEA	uCi/mL	100.8	<1.50e-04	< 6.075	<6.08e0	N/A	N/A	N/A	6.075	N/A
S981002949	D	TIC by Acid/Countmetry	ug/mL	105.5	<1.50e-04	5.946	5.890	5.918	1.01	N/A	6.075	N/A
S981002949	D	Sn 117 by ICP/MS Acid Addition	ug/mL	89.00	<2.00e-06	2.126	2.130	2.128	0.00	N/A	8.10e-02	N/A
S981002949	D	Cs 133 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	2.049	2.100	2.075	2.41	N/A	8.10e-02	N/A
S981002949	D	Cs 135 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	2.049	2.100	2.075	2.41	N/A	9.30e-02	N/A
S981002949	A	Protactinium-231 by ICP/MS	ug/mL	N/A	<2.27e-06	<9.26e-02	<9.26e-2	N/A	N/A	N/A	1.33e-01	N/A
S981002949	A	Thorium-230 by ICP/MS	ug/mL	N/A	<3.67e-06	<1.33e-01	<1.33e-1	N/A	N/A	N/A	1.49e-01	N/A
S981002949	A	Thorium-232 by ICP/MS	ug/mL	107.0	<3.44e-06	<1.39e-01	<1.49e-1	N/A	N/A	94.00	1.39e-01	N/A
S981002949	A	Thorium-232 by ICP/MS	ug/mL	108.4	<3.66e-06	<1.48e-01	<1.48e-1	N/A	N/A	99.50	1.48e-01	N/A
S981002949	A	Pu/Am-241 by ICP/MS	ug/mL	N/A	<3.26e-06	<1.32e-01	<1.32e-1	N/A	N/A	N/A	1.32e-01	N/A
S981002949	A	Am 243/Cm 243 by ICP/MS	ug/mL	N/A	<2.71e-06	<1.10e-01	<1.10e-1	N/A	N/A	N/A	1.10e-01	N/A
S981002949	A	Pu 244/Cm 244 by ICP/MS	ug/mL	N/A	<2.71e-06	<1.10e-01	<1.10e-1	N/A	N/A	N/A	1.10e-01	N/A

CORE NUMBER: 251
SEGMENT #: 2

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

SEGMENT POSITION: L Lower Half of Segment

Sample#	RA#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err %
S981002986	F	Silver - ICP-Fusion	ug/g	100.4	<1.00e-02	<2.08e+02	<2.07e3	n/a	n/a	92.20	208.0	n/a
S981002986	F	Ar-senic - ICP-Fusion	ug/g	101.4	<1.00e-01	<2.08e+02	<2.07e3	n/a	n/a	92.20	208.0	n/a
S981002986	F	Beryllium - ICP-Fusion	ug/g	101.6	<5.00e-03	<1.04e+02	<1.03e2	n/a	n/a	98.90	104.0	n/a
S981002986	F	Cadmium - ICP-Fusion	ug/g	102.8	<5.00e-03	<1.04e+02	<1.03e2	n/a	n/a	98.90	104.0	n/a
S981002986	F	Cerium - ICP-Fusion	ug/g	100.6	<1.00e-01	<2.08e+02	<2.07e3	n/a	n/a	97.20	208.0	n/a
S981002986	F	Cobalt - ICP-Fusion	ug/g	100.2	<1.00e-02	<4.16e+02	<4.13e2	n/a	n/a	99.50	414.0	n/a
S981002986	F	Copper - ICP-Fusion	ug/g	94.80	<1.00e-02	<2.08e+02	<2.07e3	n/a	n/a	97.20	208.0	n/a
S981002986	F	Lanthanum - ICP-Fusion	ug/g	100.8	<5.00e-02	<1.04e+03	<1.03e3	n/a	n/a	96.50	104.0	n/a
S981002986	F	Magnesium - ICP-Fusion	ug/g	104.4	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	105.0	208.0	n/a
S981002986	F	Molybdenum - ICP-Fusion	ug/g	102.0	<5.00e-02	<1.04e+03	<1.03e3	n/a	n/a	107.0	104.0	n/a
S981002986	F	Nickel - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	101.0	208.0	n/a
S981002986	F	Phosphorus - ICP-Fusion	ug/g	102.8	<2.00e-01	<4.16e+03	<4.13e3	n/a	n/a	109.0	414.0	n/a
S981002986	F	Sulfur - ICP-Fusion	ug/g	101.4	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	109.0	414.0	n/a
S981002986	F	Antimony - ICP-Fusion	ug/g	99.20	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	102.0	208.0	n/a
S981002986	F	Selenium - ICP-Fusion	ug/g	102.6	<6.00e-02	<1.25e+03	<1.24e3	n/a	n/a	102.0	208.0	n/a
S981002986	F	Strontium - ICP-Fusion	ug/g	97.20	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	102.0	208.0	n/a
S981002986	F	Titanium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.08e+03	<2.07e3	n/a	n/a	100.0	208.0	n/a
S981002986	F	Vanadium - ICP-Fusion	ug/g	98.40	<1.00e-02	<2.08e+03	<2.07e3	n/a	n/a	94.80	208.0	n/a
S981002986	F	Zinc - ICP-Fusion	ug/g	102.2	<1.00e-02	<4.16e+02	<4.13e2	394.0	33.2	94.40	208.0	n/a
S981002986	F	Cobalt-60 by GER	uCi/g	100.6	<2.00e-01	<4.16e+03	<4.13e3	n/a	n/a	94.90	414.0	n/a
S981002990	F	Sn117 ICP/MS AC1019 LA-505159	ug/g	105.1	<8.32e-03	<9.63e+03	<1.13e-2	n/a	n/a	99.10	1.04e+03	n/a
S981002990	F	Cs133 ICP/MS AC1019 LA-505159	ug/g	83.90	7.90e-05	2.74e	2.740	2.744	0.36	n/a	5.847	n/a
S981002990	F	Cs135 ICP/MS AC1019 LA-505159	ug/g	109.5	3.90e-05	3.060	3.060	2.008	13.9	110.9	7.80e-02	n/a
S981002990	F	Cs137 ICP/MS AC1019 LA-505159	ug/g	n/a	<2.54e-05	<9.88e-02	<1.01e-1	3.210	9.35	n/a	7.80e-02	n/a
S981002990	F	Protactinium-231 by ICP/MS	ug/g	n/a	<1.38e-05	<5.39e-02	<5.32e-2	n/a	n/a	n/a	7.80e-02	n/a
S981002990	F	Thorium-230 by ICP/MS	ug/g	n/a	<2.72e-05	2.77e	2.080	2.427	28.5	98.0	2.40e-02	n/a
S981002990	F	Thorium-232 by ICP/MS	ug/g	103.0	<1.41e-05	<4.95e-02	<5.63e-2	n/a	n/a	102.3	5.00e-01	n/a
S981002990	F	Pu 243 Om 243 by ICP/MS	ug/g	101.4	<2.06e-05	<8.05e-02	<8.25e-2	n/a	n/a	n/a	5.00e-02	n/a
S981002990	F	Pu 244 Om 244 by ICP/MS	ug/g	n/a	<1.46e-05	<5.70e-02	<5.85e-2	n/a	n/a	n/a	5.00e-02	n/a
S981002990	F	Silver - ICP-Acid Digest	ug/g	92.30	<1.00e-02	<13.00	<13.40	13.20	3.03	89.0	7.760	n/a
S981002990	F	Beryllium - ICP-Acid Digest	ug/g	94.80	<1.00e-01	<77.40	<7.96e1	n/a	n/a	95.00	77.60	n/a
S981002990	F	Cadmium - ICP-Acid Digest	ug/g	98.20	<5.00e-03	<3.880	<3.98e0	n/a	n/a	97.20	3.880	n/a
S981002990	F	Cerium - ICP-Acid Digest	ug/g	95.60	<5.00e-03	<3.880	<3.98e0	n/a	n/a	89.20	77.60	n/a
S981002990	F	Cobalt - ICP-Acid Digest	ug/g	87.00	<1.00e-01	<77.40	<7.96e1	n/a	n/a	89.40	77.60	n/a
S981002990	F	Copper - ICP-Acid Digest	ug/g	53.60	<2.00e-02	<15.50	<1.59e1	n/a	n/a	82.40	15.50	n/a
S981002990	F	Lanthanum - ICP-Acid Digest	ug/g	87.60	<5.20e-02	<8.420	<7.96e0	n/a	n/a	89.00	77.60	n/a
S981002990	F	Magnesium - ICP-Acid Digest	ug/g	95.40	<5.00e-02	<38.80	<3.98e1	n/a	n/a	89.00	77.60	n/a
S981002990	F	Mol Wbdenum - ICP-Acid Digest	ug/g	95.00	<1.00e-01	<2.88e0	<286.0	287.0	0.70	89.00	77.60	n/a
S981002990	F	Nickel - ICP-Acid Digest	ug/g	95.80	<5.00e-02	<38.80	<3.98e1	n/a	n/a	91.40	77.60	n/a
S981002990	F	Neodymium - ICP-Acid Digest	ug/g	93.00	<1.00e-01	<77.40	<7.96e1	n/a	n/a	86.20	77.60	n/a

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Exp
S981002960	A	Phosphorus - ICP-Acid Digest	ug/g	91.80	< 2.00e-01	1.05e+03	4.39e+03	2.72e+03	123	156.5	97.00	155.0
S981002960	A	Lead - ICP-Acid Digest	ug/g	91.60	< 1.00e-01	< 77.60	< 7.96e1	n/a	n/a	97.00	77.60	n/a
S981002960	A	Sulfur - ICP-Acid Digest	ug/g	88.80	< 1.00e-01	2.00e+03	1.28e+03	1.64e+03	43.9	9.360	77.60	n/a
S981002960	A	Antimony - ICP-Acid Digest	ug/g	92.40	< 1.00e-02	< 66.50	< 4.78e1	n/a	n/a	87.60	46.40	n/a
S981002960	A	Selenium - ICP-Acid Digest	ug/g	89.80	< 1.00e-01	< 77.60	< 7.96e1	n/a	n/a	93.20	77.60	n/a
S981002960	A	Samarium - ICP-Acid Digest	ug/g	93.60	< 1.00e-01	< 77.60	< 7.96e1	n/a	n/a	87.20	77.60	n/a
S981002960	A	Strontium - ICP-Acid Digest	ug/g	95.40	< 1.00e-02	75.80	65.60	69.70	11.8	87.80	77.60	n/a
S981002960	A	Thallium - ICP-Acid Digest	ug/g	97.00	< 2.00e-02	1.47e+02	144.0	155.5	10.9	84.80	77.60	n/a
S981002960	A	Vanadium - ICP-Acid Digest	ug/g	88.60	< 2.00e-01	< 1.59e+02	< 1.59e+02	n/a	n/a	84.20	155.0	n/a
S981002960	V	Acetate by IC-DIONE 4000/4500	ug/g	96.00	< 2.00e-02	< 38.80	< 3.98e1	n/a	n/a	91.40	38.80	n/a
S981002960	V	Glycolate IC-Dione 4000/4500	ug/g	82.02	< 2.00e-02	4.90e+02	371.0	430.4	27.6	83.35	23.43	n/a
S981002960	V	Glycolate IC-Dione 4000/4500	ug/g	98.13	< 5.60e-02	< 65.61	< 66.0e1	n/a	n/a	98.13	65.64	n/a

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Exp
S981002961	D	Silver-ICP-Acid Dil.	ug/ml	98.20	< 1.00e-02	15.80	14.40	15.10	9.27	n/a	6.010	n/a
S981002961	D	Arsenic-ICP-Acid Dil.	ug/ml	99.20	< 1.00e-01	< 3.010	< 3.010	n/a	n/a	n/a	6.010	n/a
S981002961	D	Beryllium-ICP-Acid Dil.	ug/ml	101.6	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	n/a	3.000	n/a
S981002961	D	Cadmium-ICP-Acid Dil.	ug/ml	100.6	< 1.00e-03	< 3.010	< 3.010	n/a	n/a	n/a	3.000	n/a
S981002961	D	Cerium-ICP-Acid Dil.	ug/ml	100.0	< 1.00e-01	< 6.010	< 6.010	n/a	n/a	n/a	6.010	n/a
S981002961	D	Cobalt-ICP-Acid Dil.	ug/ml	98.20	< 2.00e-02	< 12.00	< 11.20e1	n/a	n/a	n/a	12.00	n/a
S981002961	D	Copper-ICP-Acid Dil.	ug/ml	94.60	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	n/a	6.010	n/a
S981002961	D	Lanthanum-ICP-Acid Dil.	ug/ml	102.8	< 1.00e-02	< 30.10	< 3.01e1	n/a	n/a	n/a	30.10	n/a
S981002961	D	Magnesium-ICP-Acid Dil.	ug/ml	100.2	< 1.00e-02	< 60.10	< 6.01e1	n/a	n/a	n/a	60.10	n/a
S981002961	D	Molybdenum-ICP-Acid Dil.	ug/ml	98.20	< 1.00e-01	< 31.30	< 30.30	30.80	3.25	n/a	30.10	n/a
S981002961	D	Neodymium-ICP-Acid Dil.	ug/ml	99.20	< 2.00e-01	3.58e+02	341.0	349.5	4.86	n/a	40.10	n/a
S981002961	D	Lead-ICP-Acid Dil.	ug/ml	99.20	< 1.00e-01	79.60	69.10	74.35	14.1	n/a	120.0	n/a
S981002961	D	Sulfur-ICP-Acid Dil.	ug/ml	96.80	< 1.00e-01	7.28e+02	710.0	719.0	2.50	n/a	60.10	n/a
S981002961	D	Antimony-ICP-Acid Dil.	ug/ml	100.6	< 1.00e-02	< 36.10	< 3.61e1	n/a	n/a	n/a	36.10	n/a
S981002961	D	Selenium-ICP-Acid Dil.	ug/ml	96.20	< 1.00e-01	< 60.10	< 6.01e1	n/a	n/a	n/a	60.10	n/a
S981002961	D	Samarium-ICP-Acid Dil.	ug/ml	98.20	< 1.00e-01	< 60.10	< 6.01e1	n/a	n/a	n/a	60.10	n/a
S981002961	D	Strontium-ICP-Acid Dil.	ug/ml	98.60	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	n/a	6.010	n/a
S981002961	D	Thallium-ICP-Acid Dil.	ug/ml	101.6	< 1.00e-02	< 11.20e+02	< 1.20e+02	n/a	n/a	n/a	120.0	n/a
S981002961	D	Vanadium-ICP-Acid Dil.	ug/ml	99.20	< 5.00e-02	< 30.10	< 3.01e1	n/a	n/a	n/a	30.10	n/a
S981002961	D	Acetate by IC-Dione 4000/4500	ug/ml	90.52	< 2.00e-02	9.33e+02	931.0	937.9	0.64	n/a	35.15	n/a
S981002961	D	Glycolate by IC-Dione 4500 ORGAC	ug/ml	92.59	< 2.00e-02	< 33.94	< 3.94e1	n/a	n/a	n/a	33.94	n/a
S981002961	D	Glycolate by GEA	ug/ml	110.3	< 8.53e-04	< 3.24e+03	< 3.24e+03	6.28e+03	1.43	n/a	3.00e+03	n/a
S981002961	D	TIC by Acid/Columnetry	ug/ml	100.7	< 1.100	6.33e+03	< 6.33e+03	n/a	n/a	n/a	5.000	n/a
S981002966	D	Sn 117 by ICP/MS Acid Addition	ug/ml	111.50	< 1.50e-04	< 3.045	< 3.050	n/a	n/a	n/a	3.045	n/a
S981002966	D	Cs 133 by ICP/MS Acid Addition	ug/ml	98.20	< 2.00e-06	5.532	5.517	5.517	0.72	n/a	4.10e-02	n/a
S981002966	D	Cs 135 by ICP/MS Acid Addition	ug/ml	98.20	< 2.00e-06	2.065	2.030	2.047	1.47	n/a	6.10e-02	n/a
S981002966	D	Cs 137 by ICP/MS Acid Addition	ug/ml	n/a	< 2.00e-06	2.012	1.970	1.991	2.01	n/a	9.30e-02	n/a
S981002966	D	Protactinium-231 by ICP/MS	ug/ml	n/a	< 2.99e-06	< 9.26e-02	< 9.26e-02	n/a	n/a	n/a	1.33e-01	n/a
S981002966	D	Thorium-229 by ICP/MS	ug/ml	n/a	< 2.79e-06	< 1.33e-01	< 1.32e-01	n/a	n/a	n/a	1.33e-01	n/a
S981002966	D	Thorium-230 by ICP/MS	ug/ml	107.0	< 3.67e-06	< 1.49e-01	< 1.49e-01	n/a	n/a	n/a	1.39e-01	n/a
S981002966	D	Thorium-232 by ICP/MS	ug/ml	107.0	< 3.44e-06	< 1.39e-01	< 1.39e-01	n/a	n/a	n/a	1.39e-01	n/a
S981002966	D	Pu/Am-241 by ICP/MS	ug/ml	108.4	< 3.66e-06	< 1.48e-01	< 1.48e-01	n/a	n/a	n/a	1.48e-01	n/a
S981002966	D	Am 243/cm 243 by ICP/MS	ug/ml	n/a	< 3.26e-06	< 1.32e-01	< 1.32e-01	n/a	n/a	n/a	1.32e-01	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD	% Spk Rec	Det Limit	Count	Err%
S981002966	1A	Pu 244/Cm 244 by ICP/MS	ug/mL	N/A	<2.77e-06	<1.10e-01	<1.10e-1	N/A	N/A	N/A	1.10e-01	N/A	N/A

Appendix A. Opportunistic Analyte Results.
B1-105 (2)

CORE NUMBER: 251
SEGMENT #: 3

SEGMENT PORTION: Drainable Liquid

Sample#	R/L#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002962	D	Silver-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	14.80	14.00	14.40	5.56	87.40	6.010	n/a
S981002962	D	Silver-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	6.10	46.01e1	n/a	n/a	96.30	60.10	n/a
S981002962	D	Artenic-ICP-Acid DIL	ug/mL	101.6	<1.00e-03	3.010	43.01e0	n/a	n/a	99.50	3.000	n/a
S981002962	D	Beryllium-ICP-Acid DIL	ug/mL	100.6	<1.00e-03	3.010	43.01e0	n/a	n/a	99.50	3.000	n/a
S981002962	D	Cadmium-ICP-Acid DIL	ug/mL	100.6	<1.00e-03	6.010	43.01e0	n/a	n/a	97.00	60.10	n/a
S981002962	D	Cerium-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	12.00	41.20e1	n/a	n/a	97.00	12.00	n/a
S981002962	D	Cobalt-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	6.010	46.01e1	n/a	n/a	91.50	6.010	n/a
S981002962	D	Copper-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	30.10	46.01e1	n/a	n/a	98.80	30.10	n/a
S981002962	D	Lanthanum-ICP-Acid DIL	ug/mL	102.8	<1.00e-01	30.10	46.01e1	n/a	n/a	98.80	30.10	n/a
S981002962	D	Magnesium-ICP-Acid DIL	ug/mL	100.2	<1.00e-01	30.10	46.01e1	n/a	n/a	98.80	30.10	n/a
S981002962	D	Niodymium-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	34.60	30.90	32.75	11.3	99.10	30.10	n/a
S981002962	D	Phosphorus-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	34.60	30.90	32.75	11.3	99.10	30.10	n/a
S981002962	D	Sulfur-ICP-Acid DIL	ug/mL	100.5	<1.00e-01	3.35e+02	317.0	326.0	5.52	96.60	120.0	n/a
S981002962	D	Tellurium-ICP-Acid DIL	ug/mL	98.20	<1.00e-01	7.24e+02	85.30	95.15	20.7	95.20	60.10	n/a
S981002962	D	Lead-ICP-Acid DIL	ug/mL	96.80	<1.00e-01	7.24e+02	691.0	707.5	4.66	95.30	60.10	n/a
S981002962	D	Antimony-ICP-Acid DIL	ug/mL	100.5	<1.00e-02	36.10	43.61e1	n/a	n/a	100.0	36.10	n/a
S981002962	D	Selenium-ICP-Acid DIL	ug/mL	98.20	<1.00e-01	60.10	46.01e1	n/a	n/a	110.0	60.10	n/a
S981002962	D	Strontium-ICP-Acid DIL	ug/mL	98.20	<1.00e-02	6.010	46.01e1	n/a	n/a	96.00	6.010	n/a
S981002962	D	Thallium-ICP-Acid DIL	ug/mL	101.6	<1.00e-02	6.010	46.01e1	n/a	n/a	98.90	6.010	n/a
S981002962	D	Thallium-ICP-Acid DIL	ug/mL	95.40	<2.00e-01	41.20e+02	41.20e2	n/a	n/a	91.90	120.0	n/a
S981002962	D	Vanadium-ICP-Acid DIL	ug/mL	97.41	<2.00e-02	2.08e+03	2.12e+03	2.09e+03	2.87	104.3	64.44	n/a
S981002962	D	Acetate by ICP/MS 4000/4500	ug/mL	100.0	<2.00e-02	62.22	46.22e1	n/a	n/a	97.40	30.10	n/a
S981002962	D	Chloride by ICP/MS 4000/4500	ug/mL	110.3	<8.53e-04	43.44e+03	43.14e3	n/a	n/a	97.40	30.10	n/a
S981002962	D	Fluoride by ICP/MS 4000/4500	ug/mL	101.2	1.000	5.62e+03	5.59e+03	5.60e+03	0.54	n/a	3.00e-03	n/a
S981002962	D	11C by Ag/AgCl/MS	ug/mL	111.7	<1.50e-04	3.045	43.05e0	n/a	n/a	97.41	3.045	n/a
S981002962	D	Sn 117 by ICP/MS Acid addition	ug/mL	98.50	<2.00e-06	2.920	6.950	6.985	1.00	100.0	4.10e-02	n/a
S981002962	D	Ca 133 by ICP/MS Acid addition	ug/mL	n/a	<2.00e-06	2.592	2.570	2.581	0.78	n/a	4.10e-02	n/a
S981002962	D	Ca 137 by ICP/MS Acid addition	ug/mL	n/a	<2.00e-06	2.491	2.610	2.550	4.71	n/a	9.30e-02	n/a
S981002962	A	Protactinium-231 by ICP/MS	ug/mL	n/a	<2.29e-06	49.26e-02	49.26e-2	n/a	n/a	n/a	9.30e-02	n/a
S981002962	A	Thorium-229 by ICP/MS	ug/mL	n/a	<3.27e-06	41.33e-01	41.32e-1	n/a	n/a	n/a	1.49e-01	n/a
S981002962	A	Thorium-230 by ICP/MS	ug/mL	n/a	<3.67e-06	41.49e-01	41.49e-1	n/a	n/a	n/a	1.49e-01	n/a
S981002962	A	Thorium-232 by ICP/MS	ug/mL	107.0	<3.44e-06	41.39e-01	41.39e-1	n/a	n/a	98.50	1.39e-01	n/a
S981002962	A	Pu 241 by ICP/MS	ug/mL	108.4	<3.66e-06	41.48e-01	41.48e-1	n/a	n/a	102.8	1.48e-01	n/a
S981002962	A	Pu 243 by ICP/MS	ug/mL	n/a	<3.22e-06	41.32e-01	41.32e-1	n/a	n/a	n/a	1.32e-01	n/a
S981002962	A	Pu 244 by ICP/MS	ug/mL	n/a	<2.71e-06	41.10e-01	41.10e-1	n/a	n/a	n/a	1.10e-01	n/a

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 251
SEGMENT #: 4

SEGMENT PORTION: L Lower Half of Segment

Sample #	RA# Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S981002987	F Silver - ICP-Fusion	ug/g	100.4	< 1.00e-02	< 2.02e-02	< 2.02e-02	N/A	N/A	91.80	202.0	N/A
S981002987	F Arsenic - ICP-Fusion	ug/g	101.4	< 1.00e-02	< 2.02e-03	< 2.02e-03	N/A	N/A	104.0	2.02e+03	N/A
S981002987	F Beryllium - ICP-Fusion	ug/g	101.6	< 5.00e-03	< 1.01e-02	< 1.01e-02	N/A	N/A	99.40	101.0	N/A
S981002987	F Cadmium - ICP-Fusion	ug/g	102.8	< 5.00e-03	< 1.01e-02	< 1.01e-02	N/A	N/A	103.0	101.0	N/A
S981002987	F Cerium - ICP-Fusion	ug/g	100.6	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	98.10	2.02e+03	N/A
S981002987	F Cobalt - ICP-Fusion	ug/g	100.2	< 1.00e-02	< 4.04e-02	< 4.04e-02	N/A	N/A	100.0	405.0	N/A
S981002987	F Copper - ICP-Fusion	ug/g	94.80	< 1.00e-02	< 2.02e-02	< 2.02e-02	N/A	N/A	97.00	202.0	N/A
S981002987	F Lanthanum - ICP-Fusion	ug/g	100.8	< 5.00e-02	< 1.01e-03	< 1.01e-03	N/A	N/A	97.00	1.01e+03	N/A
S981002987	F Manganese - ICP-Fusion	ug/g	104.4	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	102.0	2.02e+03	N/A
S981002987	F Molybdenum - ICP-Fusion	ug/g	102.0	< 5.00e-02	< 1.01e-03	< 1.01e-03	N/A	N/A	104.0	1.01e+03	N/A
S981002987	F Neodymium - ICP-Fusion	ug/g	98.40	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	94.10	2.02e+03	N/A
S981002987	F Phosphorus - ICP-Fusion	ug/g	102.8	< 2.00e-01	< 5.37e-03	< 5.37e-03	N/A	N/A	103.0	4.05e+03	N/A
S981002987	F Lead - ICP-Fusion	ug/g	101.4	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	102.0	2.02e+03	N/A
S981002987	F Sulfur - ICP-Fusion	ug/g	99.20	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	110.0	2.02e+03	N/A
S981002987	F Antimony - ICP-Fusion	ug/g	102.6	< 5.00e-02	< 1.21e-03	< 1.21e-03	N/A	N/A	103.0	1.21e+03	N/A
S981002987	F Selenium - ICP-Fusion	ug/g	97.20	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	100.0	2.02e+03	N/A
S981002987	F Strontium - ICP-Fusion	ug/g	98.00	< 1.00e-01	< 2.02e-03	< 2.02e-03	N/A	N/A	95.50	2.02e+03	N/A
S981002987	F Vanadium - ICP-Fusion	ug/g	98.40	< 1.00e-02	< 2.02e-02	< 2.02e-02	N/A	N/A	99.80	202.0	N/A
S981002987	F Thallium - ICP-Fusion	ug/g	102.2	< 1.00e-02	< 3.66e-02	< 3.66e-02	N/A	N/A	99.50	4.05e+03	N/A
S981002987	F Tin - ICP-Fusion	ug/g	100.6	< 5.00e-02	< 1.01e-03	< 1.01e-03	N/A	N/A	99.50	1.01e+03	N/A
S981002987	F Vanadium - ICP-Fusion	ug/g	105.8	< 1.35e-02	< 1.47e-02	< 1.47e-02	N/A	N/A	99.50	1.50e-02	N/A
S981002987	F Cobalt-60 by GEA	ug/g	83.90	7.00e-03	< 5.977	< 5.977	N/A	N/A	N/A	5.978	N/A
S981002991	A Sn117 ICP/MS Acid Dig LA-505150	ug/g	109.5	9.90e-05	2.250	2.250	N/A	N/A	113.4	8.00e-02	N/A
S981002991	A Ca133 ICP/MS Acid Dig LA-505150	ug/g	109.5	3.90e-05	1.494	1.480	N/A	N/A	8.00e-02	N/A	N/A
S981002991	A Ca135 ICP/MS Acid Dig LA-505150	ug/g	N/A	3.90e-05	2.423	2.450	N/A	N/A	8.00e-02	N/A	N/A
S981002991	A Ca137 ICP/MS Acid Dig LA-505150	ug/g	N/A	3.90e-05	< 1.01e-01	< 1.03e-1	N/A	N/A	N/A	1.01e-01	N/A
S981002991	A Protactinium-231 by ICP/MS	ug/g	N/A	< 2.54e-05	< 1.01e-01	< 1.03e-1	N/A	N/A	N/A	5.50e-02	N/A
S981002991	A Thorium-229 by ICP/MS	ug/g	N/A	< 2.54e-05	< 5.51e-02	< 5.64e-2	N/A	N/A	95.50	5.60e-02	N/A
S981002991	A Thorium-230 by ICP/MS	ug/g	103.0	< 2.72e-05	3.072	2.700	N/A	N/A	99.42	5.60e-02	N/A
S981002991	A Thorium-232 by ICP/MS	ug/g	101.4	< 1.41e-05	< 5.61e-02	< 5.75e-2	N/A	N/A	99.42	5.60e-02	N/A
S981002991	A PuAm-241 by ICP/MS	ug/g	N/A	< 2.06e-05	< 8.43e-02	< 8.43e-2	N/A	N/A	N/A	5.80e-02	N/A
S981002991	A Pu 244/Am 244 by ICP/MS	ug/g	N/A	< 1.44e-05	< 8.8e-02	< 9.7e-2	N/A	N/A	N/A	5.80e-02	N/A
S981002991	A Silver - ICP-Acid Digest	ug/g	92.30	< 1.00e-02	12.00	13.10	N/A	N/A	94.50	7.920	N/A
S981002991	A Arsenic - ICP-Acid Digest	ug/g	94.00	< 1.00e-01	< 79.30	< 81.2e1	N/A	N/A	95.60	79.20	N/A
S981002991	A Beryllium - ICP-Acid Digest	ug/g	98.20	< 5.00e-03	< 3.970	< 4.06e0	N/A	N/A	93.60	3.960	N/A
S981002991	A Cadmium - ICP-Acid Digest	ug/g	92.80	< 5.00e-03	< 7.770	< 8.12e1	N/A	N/A	93.60	3.960	N/A
S981002991	A Cerium - ICP-Acid Digest	ug/g	95.60	< 1.00e-01	< 79.30	< 81.2e1	N/A	N/A	95.20	79.20	N/A
S981002991	A Cobalt - ICP-Acid Digest	ug/g	93.00	< 5.20e-02	< 15.90	< 1.62e1	N/A	N/A	92.80	15.90	N/A
S981002991	A Copper - ICP-Acid Digest	ug/g	87.60	< 5.20e-02	< 39.70	< 4.06e1	N/A	N/A	87.20	7.920	N/A
S981002991	A Lanthanum - ICP-Acid Digest	ug/g	95.00	< 1.00e-01	< 79.30	< 81.2e1	N/A	N/A	92.80	39.60	N/A
S981002991	A Magnesium - ICP-Acid Digest	ug/g	93.00	< 1.00e-01	< 1.85e+02	200.0	N/A	N/A	126.2	79.20	N/A
S981002991	A Molybdenum - ICP-Acid Digest	ug/g	95.80	< 5.00e-02	< 39.70	< 4.06e1	N/A	N/A	94.40	39.60	N/A
S981002991	A Neodymium - ICP-Acid Digest	ug/g	93.00	< 1.00e-01	< 79.30	< 81.2e1	N/A	N/A	90.20	79.20	N/A

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002961	A	Phosphorus -ICP-Acid Digest	ug/g	91.80	< 2.00e-01	3.18e+03	2.79e+03	2.98e+03	13.1	6.320	159.0	n/a
S98T002961	A	Lead -ICP-Acid Digest	ug/g	91.80	< 2.00e-01	< 79.30	< 48.12e1	n/a	n/a	100.0	79.20	n/a
S98T002961	A	Sulfur -ICP-Acid Digest	ug/g	88.80	< 2.00e-01	1.84e+03	2.14e+03	1.99e+03	15.1	104.4	79.20	n/a
S98T002961	A	Antimony -ICP-Acid Digest	ug/g	92.40	< 5.00e-02	< 47.60	< 44.87e1	n/a	n/a	93.00	47.60	n/a
S98T002961	A	Selenium -ICP-Acid Digest	ug/g	89.80	< 2.00e-01	79.30	< 48.12e1	n/a	n/a	94.40	79.20	n/a
S98T002961	A	Strontium -ICP-Acid Digest	ug/g	93.60	< 2.00e-01	79.30	< 48.12e1	n/a	n/a	91.00	79.20	n/a
S98T002961	A	Titanium -ICP-Acid Digest	ug/g	95.40	< 2.00e-02	84.50	74.6	79.65	12.2	92.60	7.920	n/a
S98T002961	A	Thallium -ICP-Acid Digest	ug/g	97.00	< 2.00e-01	< 1.59e+02	< 1.62e+02	1.69	7.00	88.60	159.0	n/a
S98T002961	A	Vanadium -ICP-Acid Digest	ug/g	88.60	< 5.00e-02	< 39.70	< 44.06e1	n/a	n/a	94.20	39.60	n/a
S98T002965	A	Acetate by IC-DIONE 4000/4500	ug/g	82.02	< 2.00e-02	4.41e+02	561.0	501.0	24.0	85.71	23.66	n/a
S98T002965	A	Glycolate by IC-DIONE 4000/4500	ug/g	98.13	< 5.60e-02	< 66.25	< 46.63e1	n/a	n/a	74.58	66.24	n/a

Drainable Liquid:

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002963	D	Silver-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-02	16.20	14.40	15.30	11.8	91.00	6.010	n/a
S98T002963	D	Arsenic-ICP-Acid Dil.	ug/mL	100.8	< 1.00e-01	< 60.10	< 46.01e1	n/a	n/a	98.10	60.10	n/a
S98T002963	D	Beryllium-ICP-Acid Dil.	ug/mL	101.6	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	100.0	3.000	n/a
S98T002963	D	Cadmium-ICP-Acid Dil.	ug/mL	102.0	< 5.00e-03	< 3.010	< 3.010	n/a	n/a	100.0	3.000	n/a
S98T002963	D	Cerium-ICP-Acid Dil.	ug/mL	100.0	< 1.00e-01	< 60.10	< 46.01e1	n/a	n/a	97.70	60.10	n/a
S98T002963	D	Cobalt-ICP-Acid Dil.	ug/mL	99.60	< 2.00e-02	< 12.00	< 1.20e1	n/a	n/a	97.60	12.00	n/a
S98T002963	D	Copper-ICP-Acid Dil.	ug/mL	94.40	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	91.50	6.010	n/a
S98T002963	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.4	< 5.00e-02	< 30.10	< 43.01e1	n/a	n/a	98.50	30.10	n/a
S98T002963	D	Magnesium-ICP-Acid Dil.	ug/mL	100.8	< 5.00e-02	< 60.10	< 46.01e1	n/a	n/a	99.00	60.10	n/a
S98T002963	D	Molybdenum-ICP-Acid Dil.	ug/mL	97.80	< 1.00e-01	< 60.10	< 46.01e1	n/a	n/a	99.10	30.10	n/a
S98T002963	D	Neodymium-ICP-Acid Dil.	ug/mL	100.6	< 2.00e-01	4.48e+02	4.22	4.35	5.98	94.20	120.0	n/a
S98T002963	D	Phosphorus-ICP-Acid Dil.	ug/mL	100.4	< 1.00e-01	97.00	84.70	90.85	13.5	95.70	60.10	n/a
S98T002963	D	Sulfur-ICP-Acid Dil.	ug/mL	97.40	< 1.00e-01	7.48e+02	693.0	720.5	7.63	85.70	60.10	n/a
S98T002963	D	Antimony-ICP-Acid Dil.	ug/mL	102.0	< 5.00e-02	< 36.10	< 3.61e1	n/a	n/a	101.0	36.10	n/a
S98T002963	D	Samarium-ICP-Acid Dil.	ug/mL	96.40	< 1.00e-01	< 63.90	< 46.01e1	n/a	n/a	97.60	60.10	n/a
S98T002963	D	Selenium-ICP-Acid Dil.	ug/mL	97.40	< 1.00e-01	< 60.10	< 46.01e1	n/a	n/a	95.20	60.10	n/a
S98T002963	D	Strontium-ICP-Acid Dil.	ug/mL	98.00	< 1.00e-02	< 6.010	< 6.010	n/a	n/a	99.10	6.010	n/a
S98T002963	D	Titanium-ICP-Acid Dil.	ug/mL	101.8	< 1.00e-02	< 1.20e+02	< 1.20e+02	n/a	n/a	89.20	120.0	n/a
S98T002963	D	Thallium-ICP-Acid Dil.	ug/mL	94.80	< 5.00e-02	< 30.10	< 3.01e1	n/a	n/a	97.40	30.10	n/a
S98T002963	D	Vanadium-ICP-Acid Dil.	ug/mL	99.80	< 2.00e-02	< 2.26e+03	< 2.26e+03	2.26e+03	0.00	91.38	35.15	n/a
S98T002963	D	Acetate by IC-DIONE 4000/4500	ug/mL	99.14	< 5.00e-02	< 33.94	< 33.94	n/a	n/a	122.2	33.94	n/a
S98T002963	D	Glycolate by IC-DIONE 4500 ORGAC	ug/mL	110.3	< 8.50e-02	< 33.94	< 33.94	n/a	n/a	n/a	3.00e-03	n/a
S98T002963	D	Glycolate-60 by GEA	ug/mL	101.2	< 1.000	3.71e+03	3.58e+03	3.64e+03	3.57	n/a	5.000	n/a
S98T002968	D	TIC by Acid/Coulometry	ug/mL	83.90	< 1.50e-04	< 6.075	< 6.08e0	n/a	n/a	n/a	6.075	n/a
S98T002968	D	Sn 117 by ICP/MS Acid Addition	ug/mL	107.0	< 6.90e-06	8.513	8.330	8.421	2.14	107.5	8.10e-02	n/a
S98T002968	D	Cs 133 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	2.616	2.490	2.553	5.09	n/a	8.10e-02	n/a
S98T002968	D	Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	< 2.00e-06	2.341	2.320	2.330	0.86	n/a	8.10e-02	n/a
S98T002968	D	Protactinium-231 by ICP/MS	ug/mL	n/a	< 2.12e-05	< 48.60e-01	< 48.60e-01	n/a	n/a	n/a	8.59e-01	n/a
S98T002968	A	Thorium-229 by ICP/MS	ug/mL	n/a	n/a	< 1.159	< 1.16e0	n/a	n/a	n/a	1.159	n/a
S98T002968	A	Thorium-230 by ICP/MS	ug/mL	n/a	n/a	< 1.039	< 1.04e0	n/a	n/a	n/a	1.039	n/a
S98T002968	A	Thorium-232 by ICP/MS	ug/mL	96.40	n/a	< 1.099	< 1.10e0	n/a	n/a	85.71	1.099	n/a
S98T002968	A	Pu/Am-241 by ICP/MS	ug/mL	0.25	< 9.45e-06	< 83.8e-01	< 83.8e-01	n/a	n/a	2.69e-01	3.83e-01	n/a
S98T002968	A	Am 243/Cm 243 by ICP/MS	ug/mL	n/a	< 1.36e-05	< 5.50e-01	< 5.50e-01	n/a	n/a	n/a	5.50e-01	n/a

Sample#	R A#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98T002968	1A	Pu 244/Cm 244 by ICP/MS	ug/mL	N/A	<1.15e-05	<4.64e-01	<4.64e-1	N/A	N/A	N/A	4.64e-01		N/A

Appendix A. Opportunistic Analyte Results.

CORE NUMBER: 251
SEGMENT #: 4A

SEGMENT PORTION: L Lower Half of Segment

ATTENTION: L Lower Half of Segment												
Sample #	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Error
9581002988	F	Silver - ICP-Fusion	ug/g	100.2	<1.0e-02	<2.06e+02	<2.05e2	n/a	n/a	n/a	206.0	n/a
9581002988	F	Arsenic - ICP-Fusion	ug/g	100.2	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Beryllium - ICP-Fusion	ug/g	101.4	<5.0e-03	<1.03e+02	<1.03e2	n/a	n/a	n/a	103.0	n/a
9581002988	F	Cadmium - ICP-Fusion	ug/g	102.4	<5.0e-03	<1.03e+02	<1.03e2	n/a	n/a	n/a	103.0	n/a
9581002988	F	Cerium - ICP-Fusion	ug/g	100.2	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Cobalt - ICP-Fusion	ug/g	100.4	<2.0e-02	<4.13e+02	<4.10e2	n/a	n/a	n/a	414.0	n/a
9581002988	F	Copper - ICP-Fusion	ug/g	94.60	<1.0e-02	<2.06e+02	<2.05e2	n/a	n/a	n/a	206.0	n/a
9581002988	F	Lanthanum - ICP-Fusion	ug/g	100.6	<5.0e-01	<2.06e+03	<1.03e3	n/a	n/a	n/a	1,03e+03	n/a
9581002988	F	Magnesium - ICP-Fusion	ug/g	103.6	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Molybdenum - ICP-Fusion	ug/g	101.8	<5.0e-02	<1.03e+03	<1.03e3	n/a	n/a	n/a	1,03e+03	n/a
9581002988	F	Neodymium - ICP-Fusion	ug/g	98.00	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Phosphorus - ICP-Fusion	ug/g	101.2	<2.0e-01	<4.13e+03	<4.10e3	n/a	n/a	n/a	4,14e+03	n/a
9581002988	F	Lead - ICP-Fusion	ug/g	101.0	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Sulfur - ICP-Fusion	ug/g	98.40	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Antimony - ICP-Fusion	ug/g	101.6	<5.0e-02	<1.24e+03	<1.23e3	n/a	n/a	n/a	1,24e+03	n/a
9581002988	F	Selenium - ICP-Fusion	ug/g	97.40	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Samarium - ICP-Fusion	ug/g	97.40	<1.0e-01	<2.06e+03	<2.05e3	n/a	n/a	n/a	2,06e+03	n/a
9581002988	F	Strontium - ICP-Fusion	ug/g	98.20	<1.0e-02	<2.06e+02	<2.05e2	n/a	n/a	n/a	206.0	n/a
9581002988	F	Titanium - ICP-Fusion	ug/g	102.2	<1.0e-02	<3.56e+02	<3.54e2	375.0	10.1	n/a	206.0	n/a
9581002988	F	Thallium - ICP-Fusion	ug/g	95.60	<2.0e-01	<4.13e+03	<4.10e3	n/a	n/a	n/a	4,14e+03	n/a
9581002988	F	Vanadium - ICP-Fusion	ug/g	100.4	<5.0e-02	<1.03e+03	<1.03e3	n/a	n/a	n/a	1,03e+03	n/a
9581002988	F	Zinc - ICP-Fusion	ug/g	104.0	<2.1e-03	<8.55e+03	<8.49e+3	n/a	n/a	n/a	9,00e+03	n/a
9581002992	G	Cobalt-40 by GC/MS	ug/g	109.5	9.90e-05	<5.891	<5.9840	n/a	n/a	n/a	5.891	n/a
9581002992	G	Sn117 ICP/MS Actidig LK-505159	ug/g	n/a	9.90e-05	1.956	1.680	1,808	14.4	n/a	7,20e-02	n/a
9581002992	G	Cs133 ICP/MS Actidig LK-505159	ug/g	n/a	9.90e-05	1.335	1.510	1,422	12.7	n/a	7,20e-02	n/a
9581002992	G	Cs137 ICP/MS Actidig LK-505159	ug/g	n/a	9.90e-05	1.335	1.510	2,662	2.26	n/a	7,20e-02	n/a
9581002992	G	Protactinium-231 by ICP/MS	ug/g	n/a	<3.54e-05	<2.635	<2.690	n/a	n/a	n/a	1,00e-01	n/a
9581002992	G	Thorium-229 by ICP/MS	ug/g	n/a	<3.54e-05	<9.94e-02	<1.01e-1	n/a	n/a	n/a	9,90e-02	n/a
9581002992	G	Thorium-230 by ICP/MS	ug/g	n/a	<3.52e-03	<9.90e-02	<9.91e-2	n/a	n/a	n/a	5,40e-02	n/a
9581002992	G	Thorium-232 by ICP/MS	ug/g	103.0	<1.35e-03	<1.34e-02	<5.91e-2	2,603	25.0	n/a	1,07e-01	n/a
9581002992	G	Pu/Am-241 by ICP/MS	ug/g	101.4	<1.41e-03	<5.35e-02	<5.61e-2	n/a	n/a	n/a	5,50e-02	n/a
9581002992	G	Am 243/Cm 243 by ICP/MS	ug/g	n/a	<1.26e-03	<4.11e-02	<4.82e-2	n/a	n/a	n/a	8,10e-02	n/a
9581002992	G	Pu 244/Cm 244 by ICP/MS	ug/g	n/a	<1.46e-03	<5.74e-02	<5.85e-2	n/a	n/a	n/a	3,70e-02	n/a
9581002992	G	Silver - ICP-Acid Digest	ug/g	92.30	<1.0e-02	<14.50	<12.00	13.25	18.9	n/a	7,800	n/a
9581002992	G	Arsenic - ICP-Acid Digest	ug/g	94.00	<1.0e-01	<7.20	<7.95e1	n/a	n/a	n/a	78.00	n/a
9581002992	G	Beryllium - ICP-Acid Digest	ug/g	98.20	<5.0e-03	<3.910	<3.97e0	n/a	n/a	n/a	3.910	n/a
9581002992	G	Cadmium - ICP-Acid Digest	ug/g	92.80	<5.00e-03	5.950	4.370	5,150	50.3	n/a	78.00	n/a
9581002992	G	Cerium - ICP-Acid Digest	ug/g	95.60	<1.00e-01	<78.20	<7.95e1	n/a	n/a	n/a	15.60	n/a
9581002992	G	Cobalt - ICP-Acid Digest	ug/g	93.00	<2.00e-02	<15.60	<1.59e1	13.05	8.43	n/a	7,800	n/a
9581002992	G	Copper - ICP-Acid Digest	ug/g	87.60	<5.20e-02	<13.60	<12.50	n/a	n/a	n/a	39.10	n/a
9581002992	G	Lanthanum - ICP-Acid Digest	ug/g	95.40	<1.00e-02	<39.10	<3.97e1	n/a	n/a	n/a	78.00	n/a
9581002992	G	Magnesium - ICP-Acid Digest	ug/g	93.00	<1.00e-01	<3.26e+02	834.0	580.0	87.6	n/a	39.10	n/a
9581002992	G	Molybdenum - ICP-Acid Digest	ug/g	95.80	<5.00e-02	<39.10	<3.97e1	n/a	n/a	n/a	78.00	n/a
9581002992	G	Neodymium - ICP-Acid Digest	ug/g	93.00	<1.00e-01	<78.20	<7.95e1	n/a	n/a	n/a	39.10	n/a

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002992	A	Phosphorus -ICP-Acid Digest	ug/g	91.80	< 2.00e-01	3.40e+03	2.52e+03	2.96e+03	29.7	n/a	156.0	n/a
S98T002992	A	Lead -ICP-Acid Digest	ug/g	91.60	< 1.00e-01	< 78.20	< 7.93e1	n/a	n/a	n/a	78.00	n/a
S98T002992	A	Sulfur -ICP-Acid Digest	ug/g	88.80	< 1.00e-01	1.59e+03	1.63e+03	1.61e+03	2.48	n/a	78.00	n/a
S98T002992	A	Antimony -ICP-Acid Digest	ug/g	92.40	< 6.00e-02	< 46.90	< 4.76e1	n/a	n/a	n/a	46.80	n/a
S98T002992	A	Selenium -ICP-Acid Digest	ug/g	89.80	< 1.00e-01	< 78.20	< 7.93e1	n/a	n/a	n/a	78.00	n/a
S98T002992	A	Samarium -ICP-Acid Digest	ug/g	93.60	< 1.00e-01	< 78.20	< 7.93e1	n/a	n/a	n/a	78.00	n/a
S98T002992	A	Strontium -ICP-Acid Digest	ug/g	95.40	< 1.00e-02	60.10	50.10	55.10	18.1	n/a	7.800	n/a
S98T002992	A	Titanium -ICP-Acid Digest	ug/g	97.00	< 1.00e-02	2.32e+02	233.0	232.5	0.43	n/a	7.800	n/a
S98T002992	A	Thallium -ICP-Acid Digest	ug/g	88.60	< 2.00e-01	< 1.56e+02	< 1.59e2	n/a	n/a	n/a	156.0	n/a
S98T002992	A	Vanadium -ICP-Acid Digest	ug/g	96.00	< 5.00e-02	< 39.10	< 3.97e1	n/a	n/a	n/a	39.10	n/a
S98T002996	W	Acetate by IC-DIONEX 4000/4500	ug/g	91.13	< 2.00e-02	2.29e+02	209.0	219.1	9.13	n/a	23.76	n/a
S98T002996	W	Glycolate IC-Dionex 4000/4500	ug/g	98.13	< 5.60e-02	< 66.52	< 46.59e1	n/a	n/a	n/a	66.54	n/a

Appendix A. Opportunistic Analyte Results.
B1-105 (32)

CORE NUMBER: 251
SEGMENT #: 5

SEGMENT PORTION: L Lower Half of Segment

Sample#	R#	Analyte	Unit	Standard %	Blank	Duplicate	Average	RPD %	Spk %	Det	Count	Err%
S981002989	F	Silver - ICP-Fusion	ug/g	100.2	<1.00e-02	<2.03e-02	N/A	N/A	N/A	N/A	203.0	N/A
S981002989	F	Arsenic - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Beryllium - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.03e-02	N/A	N/A	N/A	N/A	102.0	N/A
S981002989	F	Cadmium - ICP-Fusion	ug/g	102.4	<1.00e-03	<1.02e-03	N/A	N/A	N/A	N/A	102.0	N/A
S981002989	F	Cerium - ICP-Fusion	ug/g	100.2	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Cobalt - ICP-Fusion	ug/g	100.2	<1.00e-02	<4.06e-02	N/A	N/A	N/A	N/A	406.0	N/A
S981002989	F	Copper - ICP-Fusion	ug/g	94.60	<1.00e-02	<2.03e-02	N/A	N/A	N/A	N/A	203.0	N/A
S981002989	F	Lanthanum - ICP-Fusion	ug/g	100.6	<1.00e-02	<1.02e-03	N/A	N/A	N/A	N/A	1.02e+03	N/A
S981002989	F	Magnesium - ICP-Fusion	ug/g	103.4	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Molybdenum - ICP-Fusion	ug/g	101.8	<1.00e-02	<1.02e-03	N/A	N/A	N/A	N/A	1.02e+03	N/A
S981002989	F	Neodymium - ICP-Fusion	ug/g	98.00	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Phosphorus - ICP-Fusion	ug/g	101.2	<2.00e-01	<4.06e-03	N/A	N/A	N/A	N/A	4.06e+03	N/A
S981002989	F	Lead - ICP-Fusion	ug/g	101.6	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Sulfur - ICP-Fusion	ug/g	98.40	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Antimony - ICP-Fusion	ug/g	101.6	<6.00e-02	<1.22e-03	N/A	N/A	N/A	N/A	203.0	N/A
S981002989	F	Selenium - ICP-Fusion	ug/g	97.40	<1.00e-01	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Strontium - ICP-Fusion	ug/g	98.30	<1.00e-02	<2.03e-03	N/A	N/A	N/A	N/A	2.03e+03	N/A
S981002989	F	Titanium - ICP-Fusion	ug/g	102.2	<1.00e-02	<2.13e-02	N/A	N/A	N/A	N/A	203.0	N/A
S981002989	F	Thallium - ICP-Fusion	ug/g	95.40	<2.00e-01	<4.06e-03	N/A	N/A	N/A	N/A	4.06e+03	N/A
S981002989	F	Vanadium - ICP-Fusion	ug/g	100.4	<5.00e-02	<1.02e-03	N/A	N/A	N/A	N/A	1.02e+03	N/A
S981002989	F	Cobalt-60 by GER	ug/g	83.90	7.00e-03	<6.013	N/A	N/A	N/A	N/A	6.014	N/A
S981002993	A	Strontium-90 by GER	ug/g	109.5	9.90e-05	<1.868	N/A	N/A	N/A	N/A	8.00e-02	N/A
S981002993	A	Gamma-137 by GER	ug/g	109.5	9.90e-05	<1.868	N/A	N/A	N/A	N/A	8.00e-02	N/A
S981002993	A	Gamma-137 by GER	ug/g	109.5	9.90e-05	<1.868	N/A	N/A	N/A	N/A	8.00e-02	N/A
S981002993	A	Gamma-137 by GER	ug/g	109.5	9.90e-05	<1.868	N/A	N/A	N/A	N/A	8.00e-02	N/A
S981002993	A	Gamma-137 by GER	ug/g	109.5	9.90e-05	<1.868	N/A	N/A	N/A	N/A	8.00e-02	N/A
S981002993	A	Protactinium-231 by CP/MS	ug/g	N/A	5.30e-05	<1.567	N/A	N/A	N/A	N/A	1.02e-01	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	N/A	<2.54e-05	<1.05e-1	N/A	N/A	N/A	N/A	5.50e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	N/A	<1.58e-05	<1.04e-1	N/A	N/A	N/A	N/A	5.50e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	N/A	<1.58e-05	<1.04e-1	N/A	N/A	N/A	N/A	5.50e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	103.0	<2.71e-05	<1.325	N/A	N/A	N/A	N/A	1.02e-01	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	101.4	<1.41e-05	<5.81e-2	N/A	N/A	N/A	N/A	5.81e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	N/A	<2.04e-05	<8.27e-02	N/A	N/A	N/A	N/A	8.30e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	N/A	<1.44e-05	<8.08e-02	N/A	N/A	N/A	N/A	5.90e-02	N/A
S981002993	A	Thorium-230 by CP/MS	ug/g	92.30	<1.00e-02	<12.70	N/A	N/A	N/A	N/A	7.960	N/A
S981002993	A	Silver - ICP-Acid Digest	ug/g	94.00	<1.00e-01	<79.80	N/A	N/A	N/A	N/A	96.00	N/A
S981002993	A	Arsenic - ICP-Acid Digest	ug/g	98.20	<5.00e-03	<3.990	N/A	N/A	N/A	N/A	90.20	N/A
S981002993	A	Beryllium - ICP-Acid Digest	ug/g	92.80	<5.00e-03	<3.990	N/A	N/A	N/A	N/A	92.00	N/A
S981002993	A	Cadmium - ICP-Acid Digest	ug/g	95.60	<1.00e-01	<79.80	N/A	N/A	N/A	N/A	89.00	N/A
S981002993	A	Cerium - ICP-Acid Digest	ug/g	93.00	<2.00e-02	<16.00	N/A	N/A	N/A	N/A	90.60	N/A
S981002993	A	Cobalt - ICP-Acid Digest	ug/g	87.60	<5.20e-02	<7.980	N/A	N/A	N/A	N/A	82.80	N/A
S981002993	A	Copper - ICP-Acid Digest	ug/g	95.40	<5.00e-02	<39.90	N/A	N/A	N/A	N/A	88.00	N/A
S981002993	A	Lanthanum - ICP-Acid Digest	ug/g	93.00	<1.00e-01	<1.71e+02	N/A	N/A	N/A	N/A	91.60	N/A
S981002993	A	Molybdenum - ICP-Acid Digest	ug/g	95.80	<5.00e-02	<39.90	N/A	N/A	N/A	N/A	99.90	N/A
S981002993	A	Neodymium - ICP-Acid Digest	ug/g	93.00	<1.00e-01	<79.80	N/A	N/A	N/A	N/A	85.00	N/A

5107

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002993	A	Phosphorus - ICP-Acid Digest	ug/g	91.80	<2.00e-01	1.65e+03	< 79.80	1.67e+03	4.79	119.4	160.0	N/A
S98T002993	A	Lead - ICP-Acid Digest	ug/g	88.80	<1.00e-01	< 79.80	<8.21e1	< 79.80	N/A	96.20	79.60	N/A
S98T002993	A	Sulfur - ICP-Acid Digest	ug/g	91.80	<1.00e-01	1.35e+03	1.56e+03	1.47e+03	12.2	89.00	79.60	N/A
S98T002993	A	Antimony - ICP-Acid Digest	ug/g	92.40	<6.00e-02	< 47.90	<4.95e1	< 47.90	N/A	88.20	48.00	N/A
S98T002993	A	Selenium - ICP-Acid Digest	ug/g	89.80	<1.00e-01	< 79.80	<8.21e1	< 79.80	N/A	92.40	79.60	N/A
S98T002993	A	Samarium - ICP-Acid Digest	ug/g	93.60	<1.00e-01	< 79.80	<8.21e1	< 79.80	N/A	86.80	79.60	N/A
S98T002993	A	Strontium - ICP-Acid Digest	ug/g	95.40	<1.00e-02	41.20	47.60	44.40	14.4	88.20	79.60	N/A
S98T002993	A	Titanium - ICP-Acid Digest	ug/g	97.00	<1.00e-02	91.70	89.10	90.40	2.88	89.00	79.60	N/A
S98T002993	A	Thallium - ICP-Acid Digest	ug/g	88.60	<2.00e-01	<1.68e+02	<1.64e2	<1.64e2	N/A	87.40	160.0	N/A
S98T002993	A	Vanadium - ICP-Acid Digest	ug/g	96.00	<5.00e-02	< 39.90	<4.11e1	< 39.90	N/A	92.00	39.90	N/A
S98T002997	W	Acetate by IC-DIONEX 4000/4500	ug/g	91.13	<2.00e-02	3.24e+02	322.0	322.8	0.62	N/A	23.62	N/A
S98T002997	W	Glycolate IC-Bionex 4000/4500	ug/g	98.13	<5.60e-02	< 66.14	<6.64e1	< 66.14	N/A	N/A	66.12	N/A

Drainable Liquid: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
S98T002944	D	Silver-ICP-Acid Dil.	ug/mL	99.20	<1.00e-02	16.00	15.90	15.95	0.63	N/A	6.010	N/A
S98T002944	D	Arsenic-ICP-Acid Dil.	ug/mL	101.6	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Beryllium-ICP-Acid Dil.	ug/mL	100.8	<5.00e-03	< 3.010	<3.01e0	< 3.010	N/A	N/A	3.000	N/A
S98T002944	D	Cadmium-ICP-Acid Dil.	ug/mL	102.0	<5.00e-03	< 3.010	<3.01e0	< 3.010	N/A	N/A	3.000	N/A
S98T002944	D	Cerium-ICP-Acid Dil.	ug/mL	100.0	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Cobalt-ICP-Acid Dil.	ug/mL	99.60	<2.00e-02	< 12.00	<1.20e1	< 12.00	N/A	N/A	12.00	N/A
S98T002944	D	Copper-ICP-Acid Dil.	ug/mL	94.40	<1.00e-02	< 6.010	<6.01e0	< 6.010	N/A	N/A	6.010	N/A
S98T002944	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.4	<5.00e-02	< 30.10	<3.01e1	< 30.10	N/A	N/A	30.10	N/A
S98T002944	D	Magnesium-ICP-Acid Dil.	ug/mL	103.4	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Molybdenum-ICP-Acid Dil.	ug/mL	100.8	<5.00e-02	< 36.80	<3.68e1	< 36.80	7.32	N/A	30.10	N/A
S98T002944	D	Neodymium-ICP-Acid Dil.	ug/mL	97.80	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Phosphorus-ICP-Acid Dil.	ug/mL	100.6	<2.00e-01	3.68e+02	407.0	387.5	10.1	N/A	120.0	N/A
S98T002944	D	Lead-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	92.00	84.40	88.20	8.62	N/A	60.10	N/A
S98T002944	D	Sulfur-ICP-Acid Dil.	ug/mL	102.0	<6.00e-02	< 36.10	<3.61e1	< 36.10	0.69	N/A	60.10	N/A
S98T002944	D	Antimony-ICP-Acid Dil.	ug/mL	96.40	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Selenium-ICP-Acid Dil.	ug/mL	97.40	<1.00e-01	< 60.10	<6.01e1	< 60.10	N/A	N/A	60.10	N/A
S98T002944	D	Samarium-ICP-Acid Dil.	ug/mL	98.00	<1.00e-02	< 6.010	<6.01e0	< 6.010	N/A	N/A	6.010	N/A
S98T002944	D	Strontium-ICP-Acid Dil.	ug/mL	101.8	<1.00e-02	< 6.010	<6.01e0	< 6.010	N/A	N/A	6.010	N/A
S98T002944	D	Titanium-ICP-Acid Dil.	ug/mL	94.80	<2.00e-01	<1.20e+02	<1.20e2	<1.20e2	N/A	N/A	120.0	N/A
S98T002944	D	Thallium-ICP-Acid Dil.	ug/mL	99.80	<5.00e-02	< 30.10	<3.01e1	< 30.10	N/A	N/A	30.10	N/A
S98T002944	D	Vanadium-ICP-Acid Dil.	ug/mL	100.9	<2.00e-02	< 8.2e+02	<8.2e+02	<8.2e+02	2.51	N/A	35.15	N/A
S98T002944	D	Acetate by IC-DIONEX 4000/4500	ug/mL	100.9	<2.00e-02	< 33.94	<3.39e1	< 33.94	N/A	N/A	33.94	N/A
S98T002944	D	Glycolate by IC-DIONEX 4500 ORGAC	ug/mL	110.3	<8.53e-04	<2.98e+03	<3.18e+03	<3.18e+03	19.1	N/A	3.00e-03	N/A
S98T002944	D	Cobalt-60 by GEA	ug/mL	90.53	<1.00e-01	< 6.075	<6.08e0	< 6.075	N/A	N/A	5.000	N/A
S98T002969	D	TiC by Acid/Coulometry	ug/mL	83.90	<1.50e-04	< 8.015	<7.820	< 7.918	N/A	N/A	6.075	N/A
S98T002969	D	Sn 117 by ICP/MS Acid Addition	ug/mL	107.0	<6.90e-06	< 2.406	<2.406	< 2.458	4.07	N/A	8.10e-02	N/A
S98T002969	D	Cs 133 by ICP/MS Acid Addition	ug/mL	107.0	<2.00e-06	< 2.232	<2.232	< 2.256	2.22	N/A	8.10e-02	N/A
S98T002969	D	Cs 137 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	< 2.232	<2.232	< 2.256	2.22	N/A	8.10e-02	N/A
S98T002969	D	Protactinium-231 by ICP/MS	ug/mL	N/A	<2.12e-05	<8.60e-01	<8.60e-01	< 8.60e-01	N/A	N/A	8.59e-01	N/A
S98T002969	A	Thorium-229 by ICP/MS	ug/mL	N/A	<1.159	<1.159	<1.159	< 1.159	N/A	N/A	1.159	N/A
S98T002969	A	Thorium-230 by ICP/MS	ug/mL	N/A	<1.039	<1.039	<1.039	< 1.039	N/A	N/A	1.039	N/A
S98T002969	A	Thorium-232 by ICP/MS	ug/mL	96.40	<1.039	<1.039	<1.039	< 1.039	N/A	N/A	1.039	N/A
S98T002969	A	Pu/Am-241 by ICP/MS	ug/mL	0.25	<9.45e-06	<3.83e-01	<3.83e-01	< 3.83e-01	N/A	N/A	3.83e-01	N/A
S98T002969	A	Am 243/Am 243 by ICP/MS	ug/mL	N/A	<1.36e-05	<5.50e-01	<5.50e-01	< 5.50e-01	N/A	N/A	5.50e-01	N/A

25-mar-1999 11:20:16
A-0002-1

Sample#	RI#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count Err%
SS81002969	A	Pu 244/Cm 244 by ICP/MS	ug/mL	n/a	<1.15e-05	<4.64e-01	<4.64e-1	n/a	n/a	n/a	4.64e-01	n/a

5109

25-mar-1999 11:20:22
A-0002-1

CORE NUMBER: 252
SEGMENT #: 5

Appendix A. Opportunistic Analyte Results.
81-105 (2)

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD % Spk Rec %	Det Limit	Count Err%
S98T003006	D	Silver-ICP-Acid Dil.	ug/mL	99.40	< 1.00e-02	13.90	14.00	13.95	0.72	n/a	n/a
S98T003006	D	Silver-ICP-Acid Dil.	ug/mL	99.60	< 1.00e-01	< 3.00	< 6.01e1	n/a	n/a	n/a	6.010
S98T003006	D	Arsenic-ICP-Acid Dil.	ug/mL	99.60	< 5.00e-03	< 3.00	< 3.01e0	n/a	n/a	n/a	6.010
S98T003006	D	Beryllium-ICP-Acid Dil.	ug/mL	101.14	< 5.00e-03	< 3.00	< 3.01e0	n/a	n/a	n/a	3.000
S98T003006	D	Cadmium-ICP-Acid Dil.	ug/mL	99.20	< 1.00e-01	< 6.00	< 6.01e1	n/a	n/a	n/a	6.010
S98T003006	D	Cerium-ICP-Acid Dil.	ug/mL	99.20	< 2.00e-02	< 12.00	< 1.20e1	n/a	n/a	n/a	12.000
S98T003006	D	Cobalt-ICP-Acid Dil.	ug/mL	92.40	< 1.00e-02	< 6.00	< 6.01e0	n/a	n/a	n/a	6.010
S98T003006	D	Copper-ICP-Acid Dil.	ug/mL	98.60	< 1.00e-02	< 30.00	< 3.01e1	n/a	n/a	n/a	30.10
S98T003006	D	Lanthanum-ICP-Acid Dil.	ug/mL	100.8	< 1.00e-01	< 60.00	< 6.01e1	32.40	3.09	n/a	30.10
S98T003006	D	Magnesium-ICP-Acid Dil.	ug/mL	100.8	< 1.00e-02	< 32.90	< 3.190	n/a	n/a	n/a	30.10
S98T003006	D	Molybdenum-ICP-Acid Dil.	ug/mL	96.00	< 1.00e-01	< 60.00	< 6.01e1	408.5	8.57	n/a	120.0
S98T003006	D	Nickel-ICP-Acid Dil.	ug/mL	99.40	< 1.00e-01	< 4.26e02	391.0	72.85	12.2	n/a	60.10
S98T003006	D	Phosphorus-ICP-Acid Dil.	ug/mL	99.10	< 1.00e-01	68.40	77.30	823.0	0.97	n/a	60.10
S98T003006	D	Lead-ICP-Acid Dil.	ug/mL	97.60	< 1.00e-01	8.27e02	819.0	n/a	n/a	n/a	36.10
S98T003006	D	Sulfur-ICP-Acid Dil.	ug/mL	99.10	< 1.00e-02	< 36.10	< 3.61e1	n/a	n/a	n/a	60.10
S98T003006	D	Antimony-ICP-Acid Dil.	ug/mL	95.40	< 1.00e-01	< 60.00	< 6.01e1	n/a	n/a	n/a	60.10
S98T003006	D	Selenium-ICP-Acid Dil.	ug/mL	96.20	< 1.00e-02	< 6.00	< 6.01e0	n/a	n/a	n/a	6.010
S98T003006	D	Samarium-ICP-Acid Dil.	ug/mL	96.20	< 1.00e-02	< 6.00	< 6.01e0	n/a	n/a	n/a	6.010
S98T003006	D	Strontium-ICP-Acid Dil.	ug/mL	100.2	< 1.00e-02	< 6.00	< 6.01e0	n/a	n/a	n/a	6.010
S98T003006	D	Titanium-ICP-Acid Dil.	ug/mL	96.60	< 2.00e-02	< 30.10	< 3.01e1	n/a	n/a	n/a	30.10
S98T003006	D	Thallium-ICP-Acid Dil.	ug/mL	99.20	< 2.00e-02	< 1.96e+03	1.99e+03	1.97e+03	1.52	n/a	123.0
S98T003006	D	Vanadium-ICP-Acid Dil.	ug/mL	96.23	< 2.00e-02	< 1.96e+02	< 1.96e+02	n/a	n/a	n/a	123.0
S98T003006	D	Acetate by IC-Dionex 4500	ug/mL	94.44	< 2.00e-02	< 1.96e+02	< 1.96e+02	n/a	n/a	n/a	3.00e-03
S98T003006	D	Glycolate-IC-Dionex 4500	ug/mL	106.9	< 2.71e-03	< 2.90e-03	< 3.05e-3	n/a	n/a	n/a	5.000
S98T003006	D	Cobalt-60 by GEA	ug/mL	100.3	< 1.50e-04	< 6.075	< 6.08e0	4.74e+03	4.01	n/a	6.075
S98T003010	D	TIC by Acid/Coulometry	ug/mL	83.90	< 6.00e-06	< 2.268	< 2.110	7.090	4.51	n/a	8.10e-02
S98T003010	D	Ca 133 by ICP/MS Acid Addition	ug/mL	107.0	< 5.00e-06	2.268	2.110	2.189	7.31	n/a	8.10e-02
S98T003010	D	Ca 135 by ICP/MS Acid Addition	ug/mL	n/a	< 5.00e-06	1.956	1.920	1.938	2.06	n/a	8.10e-02
S98T003010	D	Ca 137 by ICP/MS Acid Addition	ug/mL	n/a	< 5.12e-05	< 48.60e-01	< 48.60e-01	n/a	n/a	n/a	8.10e-02
S98T003010	D	Protactinium-231 by ICP/MS	ug/mL	n/a	< 1.159	< 1.159	< 1.159	n/a	n/a	n/a	1.159
S98T003010	D	Thorium-229 by ICP/MS	ug/mL	n/a	< 1.099	< 1.099	< 1.099	n/a	n/a	n/a	1.099
S98T003010	D	Thorium-230 by ICP/MS	ug/mL	96.40	< 4.55e-06	< 3.83e-01	< 3.83e-01	n/a	n/a	n/a	3.83e-01
S98T003010	D	Thorium-232 by ICP/MS	ug/mL	0.23	< 1.34e-05	< 5.50e-01	< 5.50e-01	n/a	n/a	n/a	5.50e-01
S98T003010	A	Pu 241 by ICP/MS	ug/mL	n/a	< 1.15e-05	< 4.64e-01	< 4.64e-01	n/a	n/a	n/a	4.64e-01
S98T003010	A	Pu 243/244 by ICP/MS	ug/mL	n/a	< 1.15e-05	< 4.64e-01	< 4.64e-01	n/a	n/a	n/a	4.64e-01
S98T003010	A	Pu 244/244 by ICP/MS	ug/mL	n/a	< 1.15e-05	< 4.64e-01	< 4.64e-01	n/a	n/a	n/a	4.64e-01

5110

CORE NUMBER: 252
SEGMENT #: 6

Appendix A. Opportunistic Analyte Results.
81-105 (2)

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicates	Average	Rep. % Spk Rec. %	Det Limit	Count	Err%
S981003007	D	Silver-ICP-Acid DIL.	ug/mL	99.40	<1.00e-02	4.70	13.70	14.20	7.04	N/A	6.010	N/A
S981003007	D	Arsenic-ICP-Acid DIL.	ug/mL	99.60	<1.00e-01	< 6.010	<6.01e1	N/A	N/A	N/A	60.10	N/A
S981003007	D	Beryllium-ICP-Acid DIL.	ug/mL	99.60	<5.00e-03	< 3.010	<3.01e0	N/A	N/A	N/A	3.000	N/A
S981003007	D	Cadmium-ICP-Acid DIL.	ug/mL	101.4	<5.00e-03	< 3.010	<3.01e0	N/A	N/A	N/A	3.000	N/A
S981003007	D	Cerium-ICP-Acid DIL.	ug/mL	98.20	<1.00e-01	< 12.010	<1.20e1	N/A	N/A	N/A	12.000	N/A
S981003007	D	Cobalt-ICP-Acid DIL.	ug/mL	99.20	<2.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S981003007	D	Copper-ICP-Acid DIL.	ug/mL	96.40	<1.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	30.10	N/A
S981003007	D	Lanthanum-ICP-Acid DIL.	ug/mL	100.8	<5.00e-02	< 32.50	<32.5e1	N/A	N/A	N/A	30.10	N/A
S981003007	D	Magnesium-ICP-Acid DIL.	ug/mL	100.8	<1.00e-01	< 32.50	<32.5e1	N/A	N/A	N/A	30.10	N/A
S981003007	D	Neodymium-ICP-Acid DIL.	ug/mL	96.00	<1.00e-01	< 33.5e02	<34.0	37.5	5.09	N/A	120.0	N/A
S981003007	D	Phosphorus-ICP-Acid DIL.	ug/mL	97.10	<2.00e-01	< 82.40	<82.40	76.95	19.9	N/A	60.10	N/A
S981003007	D	Lead-ICP-Acid DIL.	ug/mL	97.40	<1.00e-01	< 36.10	<36.1e1	811.5	1.36	N/A	60.10	N/A
S981003007	D	Selenium-ICP-Acid DIL.	ug/mL	97.40	<5.00e-02	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S981003007	D	Samarium-ICP-Acid DIL.	ug/mL	96.30	<1.00e-01	< 60.10	<6.01e1	N/A	N/A	N/A	60.10	N/A
S981003007	D	Strontium-ICP-Acid DIL.	ug/mL	100.2	<1.00e-02	< 6.010	<6.01e0	N/A	N/A	N/A	6.010	N/A
S981003007	D	Tellurium-ICP-Acid DIL.	ug/mL	96.30	<2.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	120.0	N/A
S981003007	D	Vanadium-ICP-Acid DIL.	ug/mL	96.30	<5.00e-02	< 30.10	<3.01e1	N/A	N/A	N/A	120.0	N/A
S981003007	D	Acetate By IC-Dioxex 4500	ug/mL	87.80	<2.00e-02	< 1.64e+03	<1.64e+03	1.65e+03	1.21	N/A	118.8	N/A
S981003007	D	Glycolate-IC-Dioxex 4500	ug/mL	91.40	<2.00e-02	< 3.32e+03	<2.86e+3	N/A	N/A	N/A	3.00e+03	N/A
S981003007	D	Cobalt-60 By GEA	ug/mL	100.6	<2.71e-03	< 4.57e+03	<4.57e+03	4.51e+03	2.66	N/A	5.000	N/A
S981003011	D	11C By ACP/Countercity	ug/mL	87.92	<1.50e-04	< 6.075	<6.08e0	N/A	N/A	N/A	6.075	N/A
S981003011	D	Sn 117 By ICP/MS Acid Addition	ug/mL	107.0	<1.00e-05	< 7.310	<7.310	7.440	3.49	N/A	8.10e-02	N/A
S981003011	D	Cs 135 By ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	< 2.005	<2.005	2.098	9.07	N/A	8.10e-02	N/A
S981003011	D	Cs 137 By ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	< 1.701	<1.950	1.825	13.7	N/A	8.10e-02	N/A
S981003011	D	Productium-231 By ICP/MS	ug/mL	N/A	<2.12e-05	<8.60e-01	<8.60e-1	N/A	N/A	N/A	8.59e-01	N/A
S981003011	A	Thorium-230 By ICP/MS	ug/mL	N/A	N/A	< 1.159	<1.16e0	N/A	N/A	N/A	1.159	N/A
S981003011	A	Thorium-232 By ICP/MS	ug/mL	N/A	N/A	< 1.099	<1.04e0	N/A	N/A	N/A	1.099	N/A
S981003011	A	Pa/Am-241 By ICP/MS	ug/mL	96.40	<9.45e-06	<3.83e-01	<3.83e-1	N/A	N/A	N/A	3.83e-01	N/A
S981003011	A	Am 243/cm 243 By ICP/MS	ug/mL	0.25	<1.36e-05	<5.50e-01	<5.50e-1	N/A	N/A	N/A	5.50e-01	N/A
S981003011	A	Pa 244/cm 244 By ICP/MS	ug/mL	N/A	<1.15e-05	<4.64e-01	<4.64e-1	N/A	N/A	N/A	4.64e-01	N/A

5111

Appendix A. Opportunistic Analyte Results.
81-105 (2)

CORE NUMBER: 252
SEGMENT #: 7

SEGMENT PORTION: Drainable Liquid

Sample#	R#	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	REP. % Spk	Rec. %	Det. Limit	Count	Err%
S98T003008	D	Silver-ICP-Acid DIL.	ug/mL	99.40	<1.00e-02	13.80		13.90	1.44	93.00	6.00	6,010	N/A
S98T003008	D	Arsenic-ICP-Acid DIL.	ug/mL	99.60	<1.00e-01	60.10	<6.01e1	N/A	N/A	97.20	60.10	6,010	N/A
S98T003008	D	Beryllium-ICP-Acid DIL.	ug/mL	99.80	<5.00e-03	3.010	<2.01e0	N/A	N/A	98.60	3.000	3,000	N/A
S98T003008	D	Cadmium-ICP-Acid DIL.	ug/mL	101.4	<5.00e-03	3.010	<2.01e0	N/A	N/A	100.0	3.000	3,000	N/A
S98T003008	D	Cerium-ICP-Acid DIL.	ug/mL	98.20	<1.00e-01	69.10	<1.20e1	N/A	N/A	94.90	12.00	12,000	N/A
S98T003008	D	Cobalt-ICP-Acid DIL.	ug/mL	99.20	<2.00e-02	6.010	<1.20e1	N/A	N/A	97.80	6.010	6,010	N/A
S98T003008	D	Copper-ICP-Acid DIL.	ug/mL	92.40	<1.00e-02	6.010	<2.01e1	N/A	N/A	96.70	30.10	30,100	N/A
S98T003008	D	Lanthanum-ICP-Acid DIL.	ug/mL	96.60	<5.00e-02	34.00	<6.01e1	N/A	N/A	97.20	60.10	60,100	N/A
S98T003008	D	Magnesium-ICP-Acid DIL.	ug/mL	100.8	<5.00e-01	80.10	<3.01e1	N/A	N/A	100.0	30.10	30,100	N/A
S98T003008	D	Molybdenum-ICP-Acid DIL.	ug/mL	101.0	<1.00e-01	81.70	<1.59e1	N/A	N/A	101.0	120.0	120,000	N/A
S98T003008	D	Neodymium-ICP-Acid DIL.	ug/mL	95.00	<2.00e-01	7.99e02	81.40	81.05	0.86	100.0	60.10	60,100	N/A
S98T003008	D	Phosphorus-ICP-Acid DIL.	ug/mL	101.0	<1.00e-01	7.99e02	782.0	786.0	1.02	98.40	60.10	60,100	N/A
S98T003008	D	Lead-ICP-Acid DIL.	ug/mL	101.0	<1.00e-01	81.70	<3.01e1	N/A	N/A	101.0	36.10	36,100	N/A
S98T003008	D	Sulfur-ICP-Acid DIL.	ug/mL	101.0	<5.00e-02	6.010	<6.01e1	N/A	N/A	99.20	60.10	60,100	N/A
S98T003008	D	Antimony-ICP-Acid DIL.	ug/mL	97.00	<1.00e-01	62.00	<6.01e1	N/A	N/A	93.80	6.010	6,010	N/A
S98T003008	D	Selenium-ICP-Acid DIL.	ug/mL	97.00	<1.00e-01	62.00	<6.01e1	N/A	N/A	93.80	6.010	6,010	N/A
S98T003008	D	Strontium-ICP-Acid DIL.	ug/mL	100.2	<1.00e-02	6.010	<6.01e1	N/A	N/A	97.70	6.010	6,010	N/A
S98T003008	D	Titanium-ICP-Acid DIL.	ug/mL	99.20	<2.00e-02	11.20e02	<1.20e2	N/A	N/A	97.50	30.10	30,100	N/A
S98T003008	D	Vanadium-ICP-Acid DIL.	ug/mL	99.20	<2.00e-02	1.50e+03	<1.89e+03	1.69e+03	23.0	100.8	298.8	298,800	N/A
S98T003008	D	Acetate by IC-Dioxex 4500/4500	ug/mL	97.94	<2.00e-02	<2.88e+02	<2.88e+02	1.69e+03	23.0	100.8	298.8	298,800	N/A
S98T003008	D	Glycolate-IC-Dioxex 4500 Organic	ug/mL	97.94	<2.00e-02	<2.88e+03	<3.14e+03	N/A	N/A	102.8	3.00e-03	3,000	N/A
S98T003008	D	Cobalt-60 by GeA	uCi/mL	96.85	<2.71e-03	3.23e+03	<2.93e+03	3.08e+03	9.74	88.10	5.000	5,000	N/A
S98T003012	D	TIC by GC/Coulometry	ug/mL	87.92	<1.50e-04	<6.07e-04	<6.08e0	N/A	N/A	N/A	6.07e-04	6,07e-04	N/A
S98T003012	D	Ca 133 by ICP/MS Acid Addition	ug/mL	107.0	<1.06e-05	7.930	2.270	7.800	3.33	108.5	8.10e-02	8,10e-02	N/A
S98T003012	D	Ca 135 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	2.998	2.670	2.334	5.57	N/A	8.10e-02	8,10e-02	N/A
S98T003012	D	Ca 137 by ICP/MS Acid Addition	ug/mL	N/A	<2.00e-06	1.993	1.900	1.946	4.63	N/A	5.68e-01	5,68e-01	N/A
S98T003012	D	Protectantium-231 by ICP/MS	ug/mL	N/A	<1.39e-05	<5.65e-01	<5.65e-01	N/A	N/A	N/A	1.034	1,034	N/A
S98T003012	A	Thorium-229 by ICP/MS	ug/mL	N/A	N/A	<6.74e-01	<6.74e-01	N/A	N/A	N/A	6.74e-01	6,74e-01	N/A
S98T003012	A	Thorium-230 by ICP/MS	ug/mL	106.0	N/A	<5.55e-01	<5.55e-01	N/A	N/A	89.00	5.55e-01	5,55e-01	N/A
S98T003012	A	Thorium-232 by ICP/MS	ug/mL	158.9	<2.00e-05	<8.10e-01	<8.10e-01	N/A	N/A	142.7	8.10e-01	8,10e-01	N/A
S98T003012	A	Pu 243/244 by ICP/MS	ug/mL	N/A	<1.48e-05	<5.99e-01	<5.99e-01	N/A	N/A	N/A	5.99e-01	5,99e-01	N/A
S98T003012	A	Pu 243/244 by ICP/MS	ug/mL	N/A	<8.22e-06	<3.33e-01	<3.33e-01	N/A	N/A	N/A	3.33e-01	3,33e-01	N/A

25-mar-1999 11:28:11
A-0002-1

Appendix A. Opportunistic Analyte Results.
BY-105 (2)

CORE NUMBER: 252
SEGMENT #: 8

SEGMENT PORTION: Drainable Liquid

Sample#	R #	Analyte	Unit	Standard %	Blank	Result	Duplicate	Average	RPD %	Spk Rec %	Det Limit	Count	Err%
S98T003009	D	Silver-ICP-Acid Dtl.	ug/mL	99.60	<1.00e-02	13.60	13.00	13.30	4.51	n/a	6.010	n/a	n/a
S98T003009	D	Arsenic-ICP-Acid Dtl.	ug/mL	99.60	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	60.10	n/a	n/a
S98T003009	D	Beryllium-ICP-Acid Dtl.	ug/mL	99.60	<5.00e-03	3.010	3.010e0	n/a	n/a	n/a	3.000	n/a	n/a
S98T003009	D	Cadmium-ICP-Acid Dtl.	ug/mL	101.4	<5.00e-03	3.010	3.010e0	n/a	n/a	n/a	3.000	n/a	n/a
S98T003009	D	Cerium-ICP-Acid Dtl.	ug/mL	98.20	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	12.00	n/a	n/a
S98T003009	D	Cobalt-ICP-Acid Dtl.	ug/mL	98.20	<1.00e-02	12.00	1.20e0	n/a	n/a	n/a	6.010	n/a	n/a
S98T003009	D	Copper-ICP-Acid Dtl.	ug/mL	98.60	<1.00e-02	6.010	6.010e0	n/a	n/a	n/a	30.10	n/a	n/a
S98T003009	D	Lanthanum-ICP-Acid Dtl.	ug/mL	100.6	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	60.10	n/a	n/a
S98T003009	D	Magnesium-ICP-Acid Dtl.	ug/mL	100.6	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	30.10	n/a	n/a
S98T003009	D	Molybdenum-ICP-Acid Dtl.	ug/mL	98.00	<1.00e-02	34.10	35.80	35.25	3.12	n/a	60.10	n/a	n/a
S98T003009	D	Nickelium-ICP-Acid Dtl.	ug/mL	98.00	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	60.10	n/a	n/a
S98T003009	D	Phosphorus-ICP-Acid Dtl.	ug/mL	101.0	<1.00e-01	4.49e02	154.0	451.5	1.11	n/a	120.0	n/a	n/a
S98T003009	D	Silver-ICP-Acid Dtl.	ug/mL	99.60	<1.00e-01	8.33e0	856.0	76.20	9.45	n/a	60.10	n/a	n/a
S98T003009	D	Antimony-ICP-Acid Dtl.	ug/mL	101.0	<1.00e-02	35.10	3.61e1	844.5	2.72	n/a	60.10	n/a	n/a
S98T003009	D	Selenium-ICP-Acid Dtl.	ug/mL	97.00	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	60.10	n/a	n/a
S98T003009	D	Samarium-ICP-Acid Dtl.	ug/mL	95.40	<1.00e-01	60.10	46.01e1	n/a	n/a	n/a	60.10	n/a	n/a
S98T003009	D	Strontium-ICP-Acid Dtl.	ug/mL	100.2	<1.00e-02	6.010	6.010e0	n/a	n/a	n/a	6.010	n/a	n/a
S98T003009	D	Titanium-ICP-Acid Dtl.	ug/mL	94.80	<2.00e-01	41.20e02	41.20e2	n/a	n/a	n/a	120.0	n/a	n/a
S98T003009	D	Thallium-ICP-Acid Dtl.	ug/mL	99.20	<5.00e-02	30.10	3.01e1	n/a	n/a	n/a	30.10	n/a	n/a
S98T003009	D	Vanadium-ICP-Acid Dtl.	ug/mL	97.84	<2.00e-02	1.76e03	1.59e03	1.68e03	10.1	n/a	298.8	n/a	n/a
S98T003009	D	Acetate by GC/MS	ug/mL	97.84	<2.00e-02	42.88e02	42.88e2	32.90e-3	n/a	n/a	288.5	n/a	n/a
S98T003009	D	Glycolate by GC/MS	ug/mL	106.6	<2.71e-03	3.12e-03	3.38e03	3.24e03	8.64	n/a	3.00e-03	n/a	n/a
S98T003009	D	1,6-Hexanediol by GC/MS	ug/mL	94.85	6.500	6.075	46.08e0	n/a	n/a	n/a	5.000	n/a	n/a
S98T003009	D	Sn 117 by ICP/MS Acid Addition	ug/mL	87.92	<1.50e-04	7.562	7.562	7.531	0.80	n/a	6.075	n/a	n/a
S98T003013	D	Cs 133 by ICP/MS Acid Addition	ug/mL	107.0	<2.00e-06	2.292	2.240	2.266	2.21	n/a	8.10e-02	n/a	n/a
S98T003013	D	Cs 135 by ICP/MS Acid Addition	ug/mL	n/a	<2.00e-06	1.912	1.950	1.931	2.07	n/a	8.10e-02	n/a	n/a
S98T003013	D	Cs 137 by ICP/MS Acid Addition	ug/mL	n/a	<1.39e-01	5.65e-01	5.65e-01	n/a	n/a	n/a	5.65e-01	n/a	n/a
S98T003013	A	Protactinium-231 by ICP/MS	ug/mL	n/a	n/a	<1.034	1.03e0	n/a	n/a	n/a	1.034	n/a	n/a
S98T003013	A	Thorium-229 by ICP/MS	ug/mL	n/a	n/a	<6.74e-01	6.74e-01	n/a	n/a	n/a	6.74e-01	n/a	n/a
S98T003013	A	Thorium-230 by ICP/MS	ug/mL	106.0	n/a	<5.55e-01	5.55e-01	n/a	n/a	n/a	5.55e-01	n/a	n/a
S98T003013	A	Pu Am-241 by ICP/MS	ug/mL	158.9	<2.00e-05	48.10e-01	48.10e-01	n/a	n/a	n/a	8.10e-01	n/a	n/a
S98T003013	A	Pu Am-243 by ICP/MS	ug/mL	n/a	<1.48e-01	5.99e-01	5.99e-01	n/a	n/a	n/a	5.99e-01	n/a	n/a
S98T003013	A	Pu 244/cm 244 by ICP/MS	ug/mL	n/a	<8.22e-06	43.33e-01	43.33e-01	n/a	n/a	n/a	3.33e-01	n/a	n/a