S

ENGINEERING CHANGE NOTICE

Page 1 of 2

1. ECN 657508

Proj.
ECN N/A

2. ECN Category (mark one) Supplemental ☐ Direct Revision ☒ Change ECN ☐ Temporary ☐ Standby ☐ Supersedure ☐ Cancel/Void ☐	3. Originator's Name, Organization, MSIN, and Telephone No. CE Wills, WRAP Engineering, T4-52, 373-9844		4. USQ Required? ☒ Yes ☐ No	5. Date 03/07/00
	6. Project Title/No./Work Order No. TRU Waste/101692/AJ70		7. Bldg./Sys./Fac. No. 2336-W	8. Approval Designator Q
	9. Document Numbers Changed by this ECN (Includes sheet no. and rev.) HNF-5148, Rev. 1		10. Related ECN No(s). N/A	11. Related PO No. N/A
	12a. Modification Work ☐ Yes (fill out Blk. 12b) ☒ No (NA Blks. 12b, 12c, 12d)		12b. Work Package No. N/A	12c. Modification Work Completed N/A Design Authority/Cog. Engineer Signature & Date
13a. Description of Change This document is being revised to reflect the recalculation of transmission factors required due to transmission source linkage maintenance.				
13b. Design Baseline Document? ☐ Yes ☒ No				
14a. Justification (mark one) Criteria Change ☐ Design Improvement ☐ Environmental ☐ Facility Deactivation ☐ As-Found ☒ Facilitate Const. ☐ Const. Error/Omission ☐ Design Error/Omission ☐				
14b. Justification Details Design verification not required USQ WRP-054				
15. Distribution (include name, MSIN, and no. of copies) See attached distribution sheet				
RELEASE STAMP MAR 13 2000 DATE: STA: 4 MANFORD RELEASE ID: 5				

ENGINEERING CHANGE NOTICE				Page 2 of 2	1. ECN (use no. from pg. 1) ECN-657508																																																																				
16. Design Verification Required ☐ Yes ☒ No		17. Cost Impact <table style="width: 100%;"> <tr> <td style="width: 50%; vertical-align: top;"> ENGINEERING Additional ☐ \$ N/A Savings ☐ \$ N/A </td> <td style="width: 50%; vertical-align: top;"> CONSTRUCTION Additional ☐ \$ N/A Savings ☐ \$ N/A </td> </tr> </table>			ENGINEERING Additional ☐ \$ N/A Savings ☐ \$ N/A	CONSTRUCTION Additional ☐ \$ N/A Savings ☐ \$ N/A	18. Schedule Impact (days) Improvement ☐ N/A Delay ☐ N/A																																																																		
ENGINEERING Additional ☐ \$ N/A Savings ☐ \$ N/A	CONSTRUCTION Additional ☐ \$ N/A Savings ☐ \$ N/A																																																																								
19. Change Impact Review: Indicate the related documents (other than the engineering documents identified on Side 1) that will be affected by the change described in Block 13. Enter the affected document number in Block 20.																																																																									
SDD/DD ☐ Functional Design Criteria ☐ Operating Specification ☐ Criticality Specification ☐ Conceptual Design Report ☐ Equipment Spec. ☐ Const. Spec. ☐ Procurement Spec. ☐ Vendor Information ☐ OM Manual ☐ FSAR/SAR ☐ Safety Equipment List ☐ Radiation Work Permit ☐ Environmental Impact Statement ☐ Environmental Report ☐ Environmental Permit ☐	Seismic/Stress Analysis ☐ Stress/Design Report ☐ Interface Control Drawing ☐ Calibration Procedure ☐ Installation Procedure ☐ Maintenance Procedure ☐ Engineering Procedure ☐ Operating Instruction ☐ Operating Procedure ☐ Operational Safety Requirement ☐ IEFD Drawing ☐ Cell Arrangement Drawing ☐ Essential Material Specification ☐ Fac. Proc. Samp. Schedule ☐ Inspection Plan ☐ Inventory Adjustment Request ☐	Tank Calibration Manual ☐ Health Physics Procedure ☐ Spares Multiple Unit Listing ☐ Test Procedures/Specification ☐ Component Index ☐ ASME Coded Item ☐ Human Factor Consideration ☐ Computer Software ☐ Electric Circuit Schedule ☐ ICRS Procedure ☐ Process Control Manual/Plan ☐ Process Flow Chart ☐ Purchase Requisition ☐ Tickler File ☐ <u>None</u> ☒ ☐																																																																							
20. Other Affected Documents: (NOTE: Documents listed below will not be revised by this ECN.) Signatures below indicate that the signing organization has been notified of other affected documents listed below. <table style="width: 100%;"> <tr> <td style="width: 33%; text-align: center;">Document Number/Revision</td> <td style="width: 33%; text-align: center;">Document Number/Revision</td> <td style="width: 33%; text-align: center;">Document Number/Revision</td> </tr> <tr> <td style="height: 40px; vertical-align: bottom;">None</td> <td></td> <td></td> </tr> </table>						Document Number/Revision	Document Number/Revision	Document Number/Revision	None																																																																
Document Number/Revision	Document Number/Revision	Document Number/Revision																																																																							
None																																																																									
21. Approvals <table style="width: 100%; margin-top: 10px;"> <tr> <th style="width: 45%; text-align: left;">Signature</th> <th style="width: 10%; text-align: left;">Date</th> <th style="width: 45%; text-align: left;">Signature</th> <th style="width: 10%; text-align: left;">Date</th> </tr> <tr> <td>Design Authority _____</td> <td></td> <td>Design Agent _____</td> <td></td> </tr> <tr> <td>Cog. Eng. CE Wills <u>[Signature]</u></td> <td>3/8/00</td> <td>PE _____</td> <td></td> </tr> <tr> <td>Cog. Mgr. JR Weidert <u>[Signature]</u></td> <td>3-8-00</td> <td>QA _____</td> <td></td> </tr> <tr> <td>QA WR Thackaberry <u>[Signature]</u></td> <td>03-09-2000</td> <td>Safety _____</td> <td></td> </tr> <tr> <td>Safety _____</td> <td></td> <td>Design _____</td> <td></td> </tr> <tr> <td>Environ. _____</td> <td></td> <td>Environ. _____</td> <td></td> </tr> <tr> <td>Other SPO QA. JL Maupin <u>[Signature]</u></td> <td>3/10/00</td> <td>Other _____</td> <td></td> </tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> <tr><td>_____</td><td></td><td>_____</td><td></td></tr> </table> <p style="margin-top: 20px;">DEPARTMENT OF ENERGY</p> <p>Signature or a Control Number that tracks the Approval Signature _____</p> <p style="margin-top: 20px;">ADDITIONAL _____ _____ _____ _____ _____</p>						Signature	Date	Signature	Date	Design Authority _____		Design Agent _____		Cog. Eng. CE Wills <u>[Signature]</u>	3/8/00	PE _____		Cog. Mgr. JR Weidert <u>[Signature]</u>	3-8-00	QA _____		QA WR Thackaberry <u>[Signature]</u>	03-09-2000	Safety _____		Safety _____		Design _____		Environ. _____		Environ. _____		Other SPO QA. JL Maupin <u>[Signature]</u>	3/10/00	Other _____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____		_____	
Signature	Date	Signature	Date																																																																						
Design Authority _____		Design Agent _____																																																																							
Cog. Eng. CE Wills <u>[Signature]</u>	3/8/00	PE _____																																																																							
Cog. Mgr. JR Weidert <u>[Signature]</u>	3-8-00	QA _____																																																																							
QA WR Thackaberry <u>[Signature]</u>	03-09-2000	Safety _____																																																																							
Safety _____		Design _____																																																																							
Environ. _____		Environ. _____																																																																							
Other SPO QA. JL Maupin <u>[Signature]</u>	3/10/00	Other _____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							
_____		_____																																																																							

[illegible]

S

Calibration Report For The WRAP Facility Gamma Energy Assay System

Prepared for the U.S. Department of Energy
Assistant Secretary for Environmental Management

Project Hanford Management Contractor for the
U.S. Department of Energy under Contract DE-AC06-96RL13200

Fluor Hanford

P.O. Box 1000
Richland, Washington

Calibration Report For The WRAP Facility Gamma Energy Assay System Calibration Report For The WRAP Facility

CE Wills
Fluor Hanford

EDT/ECN: ECN-657508

03/31/00
Total Pages: ~~125~~ 132

Registered Trademarks:

* Genie-PC is a registered trademark of Digital Equipment Corporation, Huston, TX.

* MGA Software is a registered trademark of Mitchell & Gauthier Associates, Inc. Concord, MA.

Date Published
March 2000

Prepared for the U.S. Department of Energy
Assistant Secretary for Environmental Management

Fluor Hanford
P.O. Box 1000
Richland, Washington

INFORMATION CLEARANCE FORM

A. Information Category ☐ Abstract ☐ Journal Article ☐ Summary ☐ Internet ☐ Visual Aid ☐ Software ☐ Full Paper ☒ Report ☐ Other _____		B. Document Number HNF-5148, Rev. 2 C. Title Calibration Report for the WRAP Facility Gamma energy Analysis System. D. Internet Address N/A	
---	--	---	--

E. Required Information 1. Is document potentially Classified? ☒ No ☐ Yes (MANDATORY) <div style="border: 1px solid black; padding: 5px; margin: 5px 0;"> Manager's Signature Required <i>[Signature]</i> ADC If Yes ☒ No ☐ Yes Classified ADC Signature Required </div> 2. Internal Review Required? ☒ No ☐ Yes If Yes, Document Signatures Below Counsel _____ Program _____ 3. References in the Information are Applied Technology ☒ No ☐ Yes Export Controlled Information ☒ No ☐ Yes		4. Does Information Contain the Following: (MANDATORY) a. New or Novel (Patentable) Subject Matter? ☒ No ☐ Yes If "Yes", Disclosure No.: _____ b. Information Received in Confidence, Such as Proprietary and/or Inventions? ☒ No ☐ Yes If "Yes", Affix Appropriate Legends/Notices. c. Copyrights? ☒ No ☐ Yes If "Yes", Attach Permission. d. Trademarks? ☐ No ☒ Yes If "Yes", Identify in Document. 5. Is Information requiring submission to OSTI? ☒ No ☐ Yes If Yes UC-_____ and B&R-_____ 6. Release Level? ☒ Public ☐ Limited 7. Charge Code <u>AJ60</u>	
--	--	--	--

F. Complete for a Journal Article
 1. Title of Journal _____

G. Complete for a Presentation
 1. Title for Conference or Meeting _____
 2. Group Sponsoring _____
 3. Date of Conference _____
 4. City/State _____
 5. Will Information be Published in Proceedings? ☐ No ☐ Yes
 6. Will Material be Handed Out? ☐ No ☐ Yes

H. Author/Requestor <u>CE Wills</u> (Print and Sign) <i>[Signature]</i>		Responsible Manager <u>JR Weidert</u> (Print and Sign) <i>[Signature]</i>	
--	--	--	--

I. Reviewers	Yes	Print	Signature	Public Y/N (If N, complete J)
General Counsel	☐	_____	_____	Y / N
Office of External Affairs	☐	_____	_____	Y / N
DOE-RL	☒	_____	_____	Y / N
Other	☐	_____	_____	Y / N
Other	☐	_____	_____	Y / N

J. If Information Includes Sensitive Information and is not to be released to the Public indicate category below.

☐ Applied Technology	☐ Protected CRADA
☐ Personal/Private	☐ Export Controlled
☐ Proprietary	☐ Procurement-Sensitive
☒ Business-Sensitive	☐ Patentable
☐ Predecisional	☐ Other (Specify) _____
☐ UCNI	

Information Clearance Approval

K. If Additional Comments, Please Attach Separate Sheet

RELEASE AUTHORIZATION

Document Number: HNF-5148, REV 2

Document Title: Calibration Report for the WRAP Facility Gamma Energy Assay System

This document, reviewed in accordance with DOE Order 241.1, "Scientific and Technical Information Management," and 241.1-1, "Guide to the Management of Scientific and Technical Information," does not contain classified or sensitive unclassified information and is:

APPROVED FOR PUBLIC RELEASE

M. A. Williams

Lockheed Martin Services, Inc.
Document Control/Information Clearance

3/13/00

Reviewed for Applied Technology, Business Sensitive, Classified, Copyrighted, Export Controlled, Patent, Personal/Private, Proprietary, Protected CRADA, Trademark, Unclassified Controlled Nuclear Information.

LEGAL DISCLAIMER. This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof. This report has been reproduced from the best available copy. Printed in the United States of America.

TRADEMARK DISCLAIMER

Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors.

This report has been reproduced from the best available copy.

Printed in the United States of America

[illegible]

CALIBRATION REPORT
FOR
GAMMA ENERGY ANALYSIS SYSTEM
104-ND-06-102 A
REVISION 1

Prepared for
Waste Receiving and Processing Facility Engineering

Prepared by
Albert I. Davis
Waste Management Technical Services

March, 2000

Certificate of Calibration

This calibration report outlines the specific items included in the TRU Control of Measuring, Testing and Data Collection Equipment⁽¹⁾.

Equipment Identifier: 104-ND-06-102 A, GEA-A

Calibration Interval: Calibration upon installation, then calibration verification annually thereafter.

Date of Calibration: 20 January, 2000 and 6 March, 2000

Calibration Verification Due date: 20 January, 2001

Limitations: 55 gallon drums.

Calibration applies to Listed equipment: 104-ND-06-102 A

Calibration Method WMH 350 Section 2.5.

Calibration range and Tolerance: Gamma energy range from 35 keV to 2.0 MeV. Tolerance error is developed within the proprietary Canberra⁽²⁾ calibration software. Analytical span, 0 to 160 grams ²³⁹Pu defined in HNF-4051⁽³⁾.

Calibration Acceptability: Calibration has been deemed acceptable with no non-conformance items.

Standards Standards are listed in Appendix B of this document.

Performers of Calibration: The system Cognizant Engineer: Craig Wills.
Signature: C. E. Wills, Date: 3/8/02

Reference to Out-of-Cal., or non-conformance issues: TRU-WRP-00CAR-045,
TRU-WRP-00CAR-046,
TRU-WRP-00CAR-047,
TRU-WRP-00CAR-048.

Table of Contents

	Page
Certificate of Calibration	2
1.0 Introduction	5
2.0 Calibration Configuration	5
3.0 GWAS Detectors	5
4.0 GWAS Configuration	6
5.0 Vertical Activity Response	7
6.0 Comparative Analysis	8
7.0 SEGe Detectors	9
8.0 Low Energy Detectors	12
9.0 SEGe Efficiency	12
10.0 Transmission Calibration	16
11.0 References	21
Appendix A, Calibration Configuration	A1-A4
Appendix B, Source Certificates	B1-B15
Appendix C, Detector Specifications	C1-C12
Appendix D, Energy & FWHM Data and Curves	D1-D8
Appendix E, SEGe Efficiency Data and Curves	E1-E51
Appendix F, Worksheet Data	F-1-F-9

Intentionally Blank

1.0 Introduction

The Waste Receiving And Processing facility (WRAP) adheres to providing gamma-ray spectroscopy instrument calibrations traceable to the National Institute for Standards and Technology (NIST) standards⁽⁴⁾. The detectors are used to produce quantitative results for the Waste Isolation Pilot Plant (WIPP) and must meet calibration programmatic calibration goals. Instruments must meet portions of ANSI N42.14, 1978 guide for Germanium detectors. The Non-Destructive Assay (NDA) Gamma Energy Analysis (GEA) utilizes NIST traceable line source standards for the detector system calibrations. The counting configuration is a series of drums containing the line sources and different density filler matrices. The drums are used to develop system efficiencies with respect to density. The efficiency and density correction factors are required for the processing of drummed waste materials of similar densities. The calibration verification is carried out after the calibration is deemed final, by counting a second drum of NIST traceable sources. Three in-depth calibrations have been completed on one of the two systems to date, the first being the system acceptance plan.

This report has a secondary function; that being the development of the instrument calibration errors which are to be folded into the Total Instrument Uncertainty document, HNF-4050.

2.0 Calibration Configuration

The two gamma systems, GEA-A and GEA-B, are calibrated utilizing a vendor supplied drum with six full length line sources distributed in a radial related pattern and differing density fillers. A study by the vendor was cited to develop the source placement. The recommended source placement pattern is considered the best representation of a homogeneous distribution of activity in the drum; see Appendix A. This drum source is the sole calibration source geometry for the Gamma Waste Analysis System, GWAS, for the development of efficiency curves. The source certifications, D846-D851, are in Appendix B. The uncertainties listed on the certificates are at 3 σ (2.576), per telecom from the supplier.

3.0 GWAS detectors

The Gamma Energy Analysis (GEA) system utilizes six detectors; four are the Segmented Energy Germanium gamma detectors (SEGe-#) used for quantitative analysis, and two Low Energy Germanium detectors (LEGe-#) which are used for peak ratios analysis of low energy gamma-rays from the 45 to 300 keV region. The segmented detectors are 22% to 25% efficient HPGe detectors (High Purity Germanium) relative to a 3x3 NaI(Tl) detector. The detector specifications are in Appendix C; they are not identified with a particular system, but, are the system detectors. The SEGe detectors have gamma attenuators that can be activated to reduce the drum gamma-ray activity to levels that the detector and their electronics can handle without signal distortion. The system drawings state the attenuators are 0.5 inch (1.125 cm) thick, and are machined Iron plate.

4.0 GWAS Configuration

The GWAS counting system⁽⁵⁾, detectors, shield and the drum elevation platform will adequately analyze 55 and 85 gallon drums (208 and 321.5 liters), see Figure 1 bottom. The drum platform elevates in two steps, 8.9 cm each, and the platform continually rotates at 10 rpm during analysis and calibrations. The middle diagram shows the SEGe gamma detector's angle of coverage for the drums. These zones of sensitivity overlap and ensure the entire drum is scanned. The lowest segment in the analysis group views only the rotational portion of the platform and is not included in the calibration. The highest segment in the group views the drum lid and "air" space, it also is not included in the calibration for 55 gallon drums.

Figure 1. NDA detector layout and zone of sensitivities.

5.0 Vertical Activity Response

A series of counts were made with the daily source check drum, following calibration, to evaluate system parameters. The system's response to the standards is shown in the next two figures. Figure 2 is a plot of the counts for selected calibration peaks versus SEGe and position. Note that analytical segment 11 is the top of the drum. The segment's count along the vertical axis is reduced for the drum extremities; the bottom two analysis groups are lower as well as the upper two analysis groups. This count distribution will reflect the efficiency curve's as well.

Figure 2.

6.0 Comparative Analysis

The daily control drum, using a separate set of six NIST traceable line sources, E838-E843, was analyzed a multiple of times to evaluate the response to the calibration. One feature was to view the activity per segment; Figure 3 portrays a plot of that activity versus segment. This plot demonstrates the system's ability to reproduce the drum's vertical isotopic activity. The plot does not indicate losses at the drum extremes. The spreadsheet for Figures 2 and 3 are in Appendix F.

Figure 3.

The following table, Table 1, displays the calibration on a second separate NIST traceable source drum. This drum is configured in the exact same geometry, but, with another second set of rod (line) sources; certificates are shown in Appendix B, E838 to E843. This particular evaluation could be considered the calibration verification.

Seg #	Seg	59 keV		81 keV		356 keV		661 keV		1332.5 keV	
		AVG	% err	AVG	% err	AVG	% err	AVG	% err	AVG	% err
11	S12	4.02	2.46	4.46	3.90	3.13	1.61	0.577	1.89	0.408	1.61
10	S13	3.73	1.44	3.58	1.04	3.08	1.61	0.547	2.08	0.373	1.61
9	S21	3.77	2.49	3.02	7.61	3.15	1.37	0.563	1.20	0.395	1.37
8	S22	3.90	1.82	3.54	6.28	3.12	1.93	0.579	1.81	0.394	1.93
7	S23	3.50	1.58	3.33	3.68	3.13	0.580	0.569	1.06	0.408	0.580
6	S31	3.61	1.67	3.45	1.43	3.14	1.39	0.601	0.688	0.384	1.39
5	S32	3.82	2.10	3.52	1.55	3.18	1.49	0.588	1.28	0.411	1.49
4	S33	3.80	2.95	3.30	1.95	3.11	1.79	0.557	1.57	0.388	1.79
3	S41	3.45	3.00	3.13	16.06	3.16	1.66	0.562	0.701	0.400	1.66
2	S42	3.54	2.70	3.59	10.82	3.22	2.64	0.566	1.56	0.366	2.64
A04		37.27	1.05	34.28	0.496	32.11	0.455	5.93	0.186	4.04	0.455
Sum	% Diff	37.14	0.35	34.91	-1.85	31.44	2.10	5.71	3.79	3.93	2.76
Src Act		35.97	2.77	31.62	2.91	31.62	2.91	5.57	3.42	3.94	3.05
Ratio	A04/Std	1.036		1.08		1.02		1.07		1.03	
	Sum/Std	1.032		1.10		0.994		1.03		0.997	

Table 1. Daily QC-57 drum response.

The total activity is calculated in Table 1, note specifically the rows for the activity ratios to the standard, per isotope. The summed spectrum analysis is contained in the A04 module and its ratio to the actual standard is A04/Std. The summation of each of the segment's activity into a summed activity is ratioed to the standard; Sum/Std. The agreement in both cases is quite good, better than 4%. The 81 keV peak of ^{133}Ba has less agreement due to the interference produced by lead X-rays from unshielded lead next to the detectors.

7.0 SEGe Detectors

The Segmented detectors are set to view energy from 35 keV to 2.0 MeV within a 4096 channel memory block. The gain and zero are carefully adjusted to ensure the resolution is not lost during the summation of the spectra of the individual segmented counts. The Canberra program also contains a built in function to aid in the zero and gain normalization for this summation. The effectiveness of this summation is demonstrated in Table 2 and its associated graph, Figure 4.

The energy equations are of the form:

$$\text{Energy (keV)} = a + b \cdot \text{chn} + c \cdot \text{chn}^2$$

Detector	Intercept Multiplier (a)	Linear Multiplier (b)	Quadratic Multiplier (c)
SEGe-1	-0.983	0.500	-1.68E-07
SEGe-2	-0.757	0.500	-1.72E-07
SEGe-3	-0.693	0.500	-1.64E-07
SEGe-4	-1.687	0.501	-1.51E-07
Summed	-1.691	0.501	1.54E-07

Table 2.

Figure 4

The Energy versus Channel plot demonstrated the good fit of the equations to the summed spectrum plot. The individual SEGe detectors are adjusted to as close as practical to the ideal setting, thus the low error values.

The equations for energy and FWHM also demonstrate this normalization effect in the consistency of the equation's coefficients as follows:

FWHM vs Energy

$$\text{FWHM(keV)} = a + b \cdot E^{0.5}$$

	a	b
SEGe-1	0.543	0.030
SEGe-2	0.378	0.039
SEGe-3	0.543	0.032
SEGe-4	0.529	0.029
Summed	0.605	0.033

Table 3.

Figure 5.

The FWHM plot does demonstrate that broadening occurs in the spectrum summation; one of the SEGe detectors (2), is wider than the others and causes the error to increase at the higher energies. The combination of these two plots suggest peak identification should be within the

acceptable parameters for the gamma analysis program. The individual curves for the detectors are with the parameter printouts in Appendix D.

8.0 Low Energy Detectors

The two LEGe HPGe detectors are 10 % efficient relative to a 3x3 NaI(Tl) detector. The detector crystal has less depth to enhance the efficiency of the lower energy gamma-rays. These detectors are shielded with 31 mil Sn. These detectors are calibrated only for energy and FWHM (Full Width at Half Maximum) parameters. The detectors need to have good resolution and have little degradation with respect to high gamma-ray throughput rates. Figure 1 shows the placement of these detectors.

The LEGe detectors, used for the MGA⁽⁶⁾ and MGAU analysis are looked upon as separate entities, that is no spectrum summing is done. The energy range for these detectors is from 35 to 300 keV stored into a 4096 channel buffer. Calibration for energy is linear and has only two terms, a requirement of the MGA/MGAU software. The equations are as follows:

$$\text{Energy (keV)} = a + b \cdot \text{Chn} \quad \text{FWHM (keV)} = a + b \cdot \text{Eng}^{0.5}$$

Detector	Energy equation		FWHM Equation	
	Intercept	Linear Multiplier	Intercept	Linear Multiplier
LEGe-5 (Upper)	0.159	0.076	0.273	0.039
LEGe-6 (Lower)	0.150	0.076	0.223	0.040

Table 4.

The FWHM for the LEGe detectors is 0.575 and 0.533 keV respectively at 60 keV and increases to 0.663 and 0.623 keV at 100 KeV, then 0.825 and 0.789 keV respectively at 200 keV. The FWHM is within the 0.85 keV resolution needed for the MGA and MGAU analysis software. The MGA and MGAU program is designed to analyze for both gamma and X-ray peak areas below 300 keV. The LEGe detectors are not calibrated for efficiency, as the Canberra program doesn't calculate activities from these spectra, only compares relative peak areas. The energy and FWHM data is included in Appendix D for the LEGe detectors.

9.0 SEGe Efficiency

The efficiency calibration output for the segmented detectors is contained in Appendix E. The efficiency curve formulas are the "best fit" selected from the regression analysis contained in the Canberra⁽⁷⁾ software. The equation's coefficients listed below are in two groups; one without the collimators (shields) activated, and the second with the collimators activated. Each equation takes the basic form:

$$\text{Ln}(\text{Eff}) = a + b \cdot \text{Ln}(E) + c \cdot \text{Ln}(E)^2 + d \cdot \text{Ln}(E)^3 + \dots$$

The spreadsheet containing the equation's coefficients are in Appendix F, F8 and F9. The efficiency parameters can regenerate portions of the Canberra's software curve, but lack the inflection point input to completely draw the efficiency curve. The error associated with each efficiency point also includes the error developed in the regression analysis matrix. The error listing or report is not readily available for printout as an equation or plot. Uncertainty for selected efficiencies is discussed below.

Figures 5 and 6 displays the efficiency curve variations with respect to density, Figure 5 has the curves for a non-shielded count, No Collimator, while Figure 6 is the efficiency curves with the shield (Collimator).

Figure 5: No Collimator

Figure 6: With Collimator

The uncertainty plots for the summed curves are seen in the following two figures, 7 and 8. Figure 7 is the plot with no shield in front of the detector, while Figure 8 has the iron shield in the analytical path.

Figure 7

Figure 8

The uncertainty for the calibration never gets less than 6.5% for the efficiency calibration error. The slight deviations from the curve in the 385 keV region is due to the low gamma yield of the upper Ba133 gamma line, longer counting would bring the efficiency error back to the base curve.

10. Transmission Calibration

The GWAS system utilizes a density evaluation method for correcting the final efficiency curves to a waste drum. The "density" correction is defined by measuring the gamma attenuation through each drum. The developed base for the correction is established during the calibration by a measure of the calibration drum with and without the shields. The program stores this data and utilizes it during actual waste drum measurements to develop the density of the waste material. The density information is then used to modify the efficiency for the waste drum and is used in the Combine All calculations. The calibration files are as follows:

Transmission Calibration Report

3-06-00 10:05:47 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE1
 Container Type: 55 gal container
 Collimator: 0

 Calibration time: 3-06-00 9:46:16 AM
 Calibrated by: System Manager
 Calibration Live Time: 123.8 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	2.94E+05	6.43E+02	4.19E+08
2	Eu-152	244.69	8.33E+04	4.03E+02	4.19E+08
3	Eu-152	344.50	2.27E+05	5.42E+02	4.19E+08
4	Eu-152	778.90	5.72E+04	3.09E+02	4.19E+08
5	Eu-152	964.10	5.53E+04	2.98E+02	4.19E+08
6	Eu-152	1112.10	4.71E+04	2.74E+02	4.19E+08
7	Eu-152	1408.00	5.90E+04	2.51E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:05:35 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE2
 Container Type: 55 gal container
 Collimator: 0

 Calibration time: 3-06-00 9:46:16 AM
 Calibrated by: System Manager
 Calibration Live Time: 195.3 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	1.90E+05	5.19E+02	4.19E+08
2	Eu-152	244.69	5.83E+04	3.22E+02	4.19E+08
3	Eu-152	344.50	1.66E+05	4.65E+02	4.19E+08
4	Eu-152	778.90	4.47E+04	2.81E+02	4.19E+08
5	Eu-152	964.10	4.31E+04	2.42E+02	4.19E+08
6	Eu-152	1112.10	3.71E+04	2.41E+02	4.19E+08
7	Eu-152	1408.00	4.68E+04	2.22E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:05:24 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE3
 Container Type: 55 gal container
 Collimator: 0

 Calibration time: 3-06-00 9:46:16 AM
 Calibrated by: System Manager
 Calibration Live Time: 178.5 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	2.12E+05	5.54E+02	4.19E+08
2	Eu-152	244.69	6.54E+04	3.46E+02	4.19E+08
3	Eu-152	344.50	1.85E+05	4.91E+02	4.19E+08
4	Eu-152	778.90	4.94E+04	3.15E+02	4.19E+08
5	Eu-152	964.10	4.82E+04	2.65E+02	4.19E+08
6	Eu-152	1112.10	4.07E+04	2.56E+02	4.19E+08
7	Eu-152	1408.00	5.20E+04	2.33E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:05:12 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE4
 Container Type: 55 gal container
 Collimator: 0

 Calibration time: 3-06-00 9:46:16 AM
 Calibrated by: System Manager
 Calibration Live Time: 240.1 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	1.24E+05	4.28E+02	4.19E+08
2	Eu-152	244.69	3.68E+04	2.75E+02	4.19E+08
3	Eu-152	344.50	1.02E+05	3.63E+02	4.19E+08
4	Eu-152	778.90	2.70E+04	2.10E+02	4.19E+08
5	Eu-152	964.10	2.60E+04	1.84E+02	4.19E+08
6	Eu-152	1112.10	2.22E+04	1.86E+02	4.19E+08
7	Eu-152	1408.00	2.85E+04	1.72E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:06:37 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE1
 Container Type: 55 gal container
 Collimator: 1

 Calibration time: 3-06-00 9:56:52 AM
 Calibrated by: System Manager
 Calibration Live Time: 232.4 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	8.48E+03	2.21E+02	4.19E+08
2	Eu-152	244.69	1.98E+04	2.50E+02	4.19E+08
3	Eu-152	344.50	7.66E+04	3.45E+02	4.19E+08
4	Eu-152	778.90	3.22E+04	2.40E+02	4.19E+08
5	Eu-152	964.10	3.49E+04	2.33E+02	4.19E+08
6	Eu-152	1112.10	3.12E+04	2.13E+02	4.19E+08
7	Eu-152	1408.00	4.42E+04	2.16E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:06:25 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE2
 Container Type: 55 gal container
 Collimator: 1

 Calibration time: 3-06-00 9:56:52 AM
 Calibrated by: System Manager
 Calibration Live Time: 265.9 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	4.46E+03	1.64E+02	4.19E+08
2	Eu-152	244.69	1.04E+04	1.95E+02	4.19E+08
3	Eu-152	344.50	4.21E+04	2.53E+02	4.19E+08
4	Eu-152	778.90	1.91E+04	1.93E+02	4.19E+08
5	Eu-152	964.10	2.04E+04	1.68E+02	4.19E+08
6	Eu-152	1112.10	1.90E+04	1.73E+02	4.19E+08
7	Eu-152	1408.00	2.69E+04	1.68E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:06:10 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE3
 Container Type: 55 gal container
 Collimator: 1

 Calibration time: 3-06-00 9:56:52 AM
 Calibrated by: System Manager
 Calibration Live Time: 259.1 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	5.00E+03	1.83E+02	4.19E+08
2	Eu-152	244.69	1.28E+04	2.06E+02	4.19E+08
3	Eu-152	344.50	5.02E+04	2.82E+02	4.19E+08
4	Eu-152	778.90	2.23E+04	2.01E+02	4.19E+08
5	Eu-152	964.10	2.45E+04	1.91E+02	4.19E+08
6	Eu-152	1112.10	2.21E+04	1.89E+02	4.19E+08
7	Eu-152	1408.00	3.11E+04	1.80E+02	4.19E+08

Transmission Calibration Report

3-06-00 10:05:59 AM

Page 1

 **** TRANSMISSION CALIBRATION REPORT ****

Detector Name: SEGE4
 Container Type: 55 gal container
 Collimator: 1

 Calibration time: 3-06-00 9:56:52 AM
 Calibrated by: System Manager
 Calibration Live Time: 281.9 seconds
 Energy Tolerance: 1.5 keV

Record	Nuclide	Energy (keV)	Peak Area (Cnts)	Peak Area Error (Cnts)	Half-Life (sec)
1	Eu-152	121.80	2.63E+03	1.14E+02	4.19E+08
2	Eu-152	244.69	5.96E+03	1.20E+02	4.19E+08
3	Eu-152	344.50	2.36E+04	1.93E+02	4.19E+08
4	Eu-152	778.90	1.02E+04	1.22E+02	4.19E+08
5	Eu-152	964.10	1.12E+04	1.33E+02	4.19E+08
6	Eu-152	1112.10	1.03E+04	1.31E+02	4.19E+08
7	Eu-152	1408.00	1.45E+04	1.22E+02	4.19E+08

REFERENCES

1. TRU Control of Measuring, Testing and Data collection Equipment, WMH-400, Waste Isolation Pilot Plant (WIPP) Procedures, Section 2.4.4, Rev 3, Sub Section 2.1.2.
2. Canberra Model S431 Genie-PC Gamma Waste Assay Software, Technical Reference Manual S431-TEC, V 2.3, May 1998.
3. Quality Assurance Objectives for Nondestructive Assay at the Waste Receiving and Processing facility, HNF-4051, Rev 3.
4. Hanford Site Transuranic Waste Certification Plan, HNF-2600, Rev 2, Appendix E.
5. WRAP Gamma Energy Assay System, KEH 5369, Installation and Maintenance Manual, 2/23/96, Document # 96048.
6. MGA: A Gamma-ray Spectrum Analysis Code for determining Plutonium isotopic abundances. Vol 1, Methods and Algorithms. NTIS-DE90013529, Lawrence Livermore National Laboratory, Livermore, Ca. April 1990.
7. Canberra Genie-PC, V2 – Advanced Topics, S400-usr, V 2.4, February 1997.

APPENDIX A

Calibration Configuration

Drum Source Distribution

Q2 SYSTEM CALIBRATION AND SOURCE DISTRIBUTION

Calibration Drum Description:

A set of four calibration matrix drums was used to determine the Q2 (Canberra GWAS) system efficiencies as a function of photon energy and source matrix density. Four 208 Liter (55 gallon) drums were homogeneously loaded with the following materials and approximate densities:

- 1) Polyurethane foam (0.14 gm/cc)
- 2) Soft Board (Homosote) (0.470 gm/cc)
- 3) Particle board (0.76 gm/cc)
- 4) Sand (1.67 g/cc)

Each calibration matrix is divided into 6 cylindrical shells, designed to simulate a homogeneous source distribution. A set of hollow tubes was mounted in each of these shells. The source tubes extended the full height of the matrix, and were mounted parallel to the center of the axis of the drum at the positions shown in the following top view diagram.

In order to simulate routine sample counting conditions during the calibration process, the drums were placed on the turntable platform and rotated at a constant speed during all the counts. The axis of rotation was perpendicular to the axis of the Q2 system detectors. In this case a line geometry source (with equal activity per unit length) positioned in a source tube provides counting results equivalent to a cylindrical shell-geometry source with the same radial distance from the center axis of the drum. Taken together, the six shells simulate a homogeneous source distribution throughout the matrix.

Each calibration drum is divided into six shells of equal volume, and labeled A, B, C, D, E, and F. The individual source tubes are numbered 1 – 9. The radial distance from the drum to the center of each of these hollow source tubes is provided in the following table.

Source Tube Location	Radius from Drum Center (inches)
1	0
2	2.3
3	5.6
4	7.3
5	8.6
6	9.8
7	10.2
8	10.5
9	10.8

Efficiency Calibration Theory

Since the cylindrical shells used in this calibration have equal volumes, count rate data obtained for calibration counts with the line sources in tubes 2, 3, 4, 5, 6, and 9 should be weighted equally for reference counting geometry efficiency determinations. Efficiency values may be calculated using the following equation:

$$\text{Efficiency} = (A2 + B3 + C4 + D5 + E6 + F9)/(\text{activity} * \text{abundance})$$

Where

Efficiency is given in the units of counts per gamma emitted from a uniform matrix, and

A2 = measured counts per second (cps) with the line source in tube 2 (shell A)

B3 = measured cps with the line source in tube 3 (shell B)

C4 = measured cps with the line source in tube 4 (shell C)

D5 = measured cps with the line source in tube 5 (shell D)

E6 = measured cps with the line source in tube 6 (shell E)

F9 = measured cps with the line source in tube 9 (shell F)

Activity = decay corrected activity of the calibration source.

APPENDIX B

Source Certificates

Calibration Certificates

The following certificates have been scanned from the originals, only one source leak page will be evident as it is a copy on the backs of each certificate. The certificates will be in two groups, one being the calibration sources and the second being the verification and daily check sources. In each set the stated uncertainty is at the 3σ level.

GWAS Calibration Source Set

A matched set of six calibrated line sources is used to perform the energy, full-width-at-half-maximum (FWHM), and the four efficiency calibrations.

Calibration Source Certificate File

The GWAS efficiency calibration requires the generation of a certificate file containing the total activity of the line sources, expressed as the number of gammas per second for each gamma line used. This is just the sum of the isotopic activities, multiplied by the abundance (branching ratio) *for each of the gamma lines*.

The current gamma rates are calculated from the current date and time and the reference date and time.

The total uncertainty for the line sources is calculated from the square root of the sum of the squares of the individual line source uncertainties (RMS).

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D846

Reference Date March 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (±)</u>	<u>Half Life⁽¹⁾</u>
Am-241	147.7 kBq (3.991 μ Ci)	3.27%	432.7 $\pm$ 0.5 years
Ba-133	206.7 kBq (5.587 μ Ci)	3.22%	10.54 $\pm$ 0.03 years
Cs-137	38.54 kBq (1.042 μ Ci)	3.07%	30.0 $\pm$ 0.2 years
Co-60	38.66 kBq (1.045 μ Ci)	2.80%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		278.388	7.09
		302.851	18.4
		355.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan

Jay Fan, Ph.D.
Calibration Laboratory Manager

March 19, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D847

Reference Date March 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (±)</u>	<u>Half Life⁽¹⁾</u>
Am-241	148.5 kBq (4.013 μ Ci)	3.25%	432.7 $\pm$ 0.5 years
Ba-133	204.9 kBq (5.538 μ Ci)	3.19%	10.54 $\pm$ 0.03 years
Cs-137	38.10 kBq (1.030 μ Ci)	2.86%	30.0 $\pm$ 0.2 years
Co-60	38.85 kBq (1.045 μ Ci)	2.83%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.857	18.4
		355.899	62.2
		383.841	8.82
Cs-137	gamma	661.860	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fen, Ph.D.
Calibration Laboratory Manager

March 19, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D848

Reference Date March 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	150.4 kBq (4.065 μ Ci)	3.25%	432.7 $\pm$ 0.5 years
Ba-133	205.4 kBq (5.552 μ Ci)	3.20%	10.54 $\pm$ 0.03 years
Cs-137	38.27 kBq (1.034 μ Ci)	2.89%	30.0 $\pm$ 0.2 years
Co-60	38.35 kBq (1.036 μ Ci)	2.66%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.851	18.4
		355.989	62.2
		383.841	8.92
Cs-137	gamma	661.860	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicity) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Farr, Ph.D.
Calibration Laboratory Manager

March 19, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D849

Reference Date March 1, 1996

Radionuclide	Activity	Total Uncertainty (%)	Half Life ⁽¹⁾
Am-241	145.1 kBq (3.920 μ Ci)	3.24%	432.7 $\pm$ 0.5 years
Ba-133	206.3 kBq (5.575 μ Ci)	3.19%	10.54 $\pm$ 0.03 years
Cs-137	37.44 kBq (1.012 μ Ci)	2.83%	30.0 $\pm$ 0.2 years
Co-60	38.30 kBq (1.035 μ Ci)	2.60%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	Type	Energy (keV)	Intensity (%)
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.851	18.4
		365.999	62.2
Cs-137	gamma	383.841	8.92
		661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

March 19, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1985.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0784

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D850

Reference Date March 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	146.4 kBq (3.956 μ Cl)	3.26%	432.7 $\pm$ 0.5 years
Ba-133	205.5 kBq (5.553 μ Cl)	3.21%	10.54 $\pm$ 0.03 years
Cs-137	37.39 kBq (1.010 μ Cl)	3.01%	30.0 $\pm$ 0.2 years
Co-60	38.52 kBq (1.041 μ Cl)	2.74%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		278.388	7.08
		302.851	18.4
		355.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.8824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

March 18, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER D851

Reference Date March 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (±)</u>	<u>Half Life⁽¹⁾</u>
Am-241	147.2 kBq (3.977 μ Ci)	3.25%	432.7 $\pm$ 0.5 years
Ba-133	205.4 kBq (5.553 μ Ci)	3.19%	10.54 $\pm$ 0.03 years
Cs-137	38.01 kBq (1.027 μ Ci)	2.88%	30.0 $\pm$ 0.2 years
Co-60	38.50 kBq (1.041 μ Ci)	2.63%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.851	18.4
		355.989	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

March 19, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

LEAK TEST METHODS

1. **Soak (Immersion) Test** (ISO/TR 4826-1979 (E) section 2.1.3)

The source is immersed in water or other suitable liquid at 50°C for at least 4 hours and the activity in the liquid measured. Acceptance limit: 0.0001 μCi alpha, 0.001 μCi beta-gamma.

2. **Immersion Test** (ANSI N542-1977, Appendix A2.1.3)

The source is immersed in water at 100°C for 10 minutes. The water is removed and the source cooled and rinsed using fresh water. These operations are repeated twice, boiling in the water from the previous rinsing operation. Acceptance limit: 0.0001 μCi alpha, 0.005 μCi beta-gamma for all the water collected.

3. **Wipe (Smear) Test** (ANSI N542-1977, Appendix A2.1.1)

All external surfaces of the source are wiped with a piece of filter paper or other absorbent material which has been moistened with an appropriate solvent and the activity removed is measured. Acceptance limit: 0.0001 μCi alpha, 0.001 μCi beta-gamma.

4. **Leak Test Not Applicable**

For sources with no covering or a delicate covering over the radioactive portion or gas standards and sources, the inactive portions or containment vessel are wipe tested as in Method 3 above. Acceptance limit: 0.0001 μCi alpha, 0.001 μCi beta-gamma.

5. **Other**

LEAK TEST CERTIFICATION

The standard was leak tested in accordance with Leak Test Method 3 as described above. The removable activity was determined to be less than 0.001 beta-gamma and 0.0001 μCi alpha. This leak test should be performed every six months or as required by a specific NRC or Agreement State License.

B. H. H. H. H.

Health Physics

March 19, 1996

Date

Calibration check and daily control drum sources.

CERTIFICATE OF CALIBRATION**MODEL ENV4057 LINE SOURCE****SERIAL NUMBER E838****Reference Date October 1, 1996**

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	221.5 kBq (5.987 μ Ci)	3.18%	432.7 $\pm$ 0.5 years
Ba-133	239.4 kBq (6.470 μ Ci)	3.34%	10.54 $\pm$ 0.03 years
Cs-137	37.20 kBq (1.005 μ Ci)	3.94%	30.0 $\pm$ 0.2 years
Co-60	37.77 kBq (1.021 μ Ci)	3.51%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		278.368	7.09
		302.851	18.4
		365.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER E839

Reference Date October 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life¹⁾</u>
Am-241	220.3 kBq (5.953 μ Ci)	3.20%	432.7 $\pm$ 0.5 years
Ba-133	242.9 kBq (6.564 μ Ci)	3.35%	10.54 $\pm$ 0.03 years
Cs-137	36.96 kBq (998.7 nCi)	4.04%	30.0 $\pm$ 0.2 years
Co-60	37.40 kBq (1.012 μ Ci)	3.60%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.851	18.4
		355.989	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.69 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1996.

● LEAK TEST CERTIFICATION ON REVERSE ●

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER E840

Reference Date October 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	222.8 kBq (6.022 μ Ci)	3.17%	432.7 $\pm$ 0.5 years
Ba-133	240.8 kBq (6.509 μ Ci)	3.32%	10.54 $\pm$ 0.03 years
Cs-137	37.08 kBq (1.002 μ Ci)	3.81%	30.0 $\pm$ 0.2 years
Co-60	36.83 kBq (995.5 nCi)	3.41%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.969	34.2
		276.388	7.09
		302.851	18.4
		355.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9824

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1996.

● LEAK TEST CERTIFICATION ON REVERSE ●

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER E841

Reference Date October 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	223.7 kBq (6.046 μ Ci)	3.18%	432.7 $\pm$ 0.5 years
Ba-133	240.6 kBq (6.503 μ Ci)	3.33%	10.54 $\pm$ 0.03 years
Cs-137	37.18 kBq (1.005 μ Ci)	3.89%	30.0 $\pm$ 0.2 years
Co-60	36.69 kBq (991.5 nCi)	3.48%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		276.388	7.09
		302.851	18.4
		355.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9624

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

● LEAK TEST CERTIFICATION ON REVERSE ●

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER E842

Reference Date October 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	225.0 kBq (6.082 μ Ci)	3.18%	432.7 $\pm$ 0.5 years
Ba-133	240.9 kBq (6.510 μ Ci)	3.33%	10.54 $\pm$ 0.03 years
Cs-137	37.13 kBq (1.003 μ Ci)	3.88%	30.0 $\pm$ 0.2 years
Co-60	37.07 kBq (1.002 μ Ci)	3.47%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.909	34.2
		276.368	7.09
		302.851	18.4
		355.999	62.2
		363.841	8.92
Cs-137	gamma	661.660	85.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9624

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.84").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.
Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1993.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

CERTIFICATE OF CALIBRATION

MODEL ENV4057 LINE SOURCE

SERIAL NUMBER E843

Reference Date October 1, 1996

<u>Radionuclide</u>	<u>Activity</u>	<u>Total Uncertainty (%)</u>	<u>Half Life⁽¹⁾</u>
Am-241	224.5 kBq (6.067 μ Ci)	3.19%	432.7 $\pm$ 0.5 years
Ba-133	241.3 kBq (6.523 μ Ci)	3.34%	10.54 $\pm$ 0.03 years
Cs-137	36.58 kBq (988.6 nCi)	3.98%	30.0 $\pm$ 0.2 years
Co-60	36.81 kBq (994.9 nCi)	3.54%	5.271 $\pm$ 0.001 years

PRINCIPLE EMISSIONS⁽¹⁾

	<u>Type</u>	<u>Energy (keV)</u>	<u>Intensity (%)</u>
Am-241	gamma	59.5364	35.7
Ba-133	gamma	80.989	34.2
		278.368	7.09
		302.851	18.4
		355.999	62.2
		383.841	8.92
Cs-137	gamma	661.660	95.21
Co-60	gamma	1173.237	99.90
		1332.501	99.9624

SOURCE DESCRIPTION

The activity is uniformly distributed in an epoxy matrix with a density of 1.07 g/cc and cast in a 9.53 mm (0.375") O.D. by 813 mm (32.0") long Aluminum tube with 0.89 mm (0.035") thick walls. 2 mm (0.080") end plugs are placed at both ends of the tube to give an active length of 809 mm (31.8").

METHOD OF CALIBRATION

The source was calibrated on a high purity Germanium detector against an NIST traceable standard in a similar geometry. The activity of the standard was determined using an energy vs. efficiency curve established through ongoing intercomparisons with the National Institute of Standards and Technology. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

Jay Fan, Ph.D.

Calibration Laboratory Manager

October 8, 1996

Date

REFERENCES

- (1) Table of Radioactive Isotopes, 7th edition, 1986.

• LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 503-9201 Fax (818) 503-0764

APPENDIX C

GWAS Detectors

Specifications

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC1018S SERIAL NUMBER 3963626CRYOSTAT MODEL 7935-7S PREAMPLIFIER MODEL 2002CSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume 45 cc Relative Efficiency 10 %Resolution 1.8 keV (FWHM) at 1.33 MeV0.70 keV (FWTM) at 1.33 MeV0.70 keV (FWHM) at 122 keV0.70 keV (FWTM) at 122 keVPeak / Compton 1 Cryostat well diameter 5 mm Cryostat well depth 5 mm

Cryostat description if special _____

Physical Characteristics

Geometry Closed-end coaxialDiameter 45.5 mmLength 80 mmDistance from window 5 mmActive Volume 45 ccWell Depth 5 mmWell Diameter 5 mm

Electrical Characteristics

Depletion voltage (+)1200 V dcRecommended bias voltage (+)3000 V dcLeakage current at recommended bias 0.01 nAPreamplifier test point voltage at recommended voltage (-)0.28 V dcCapacitance at recommended bias ~9 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.69	1.74			
FWTM (KeV)	1.27	3.25			
Peak / Compton		38.0:1			
Rel. Efficiency %		10.0			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.5 Liters per Day.Tested by: Peter Gilman Date: March 11, 1996Approved by: Margaret Lindstedt Date: March 11, 1996

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

Detector assigned to GEA-A, LEGe 6.

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC1018/S

SERIAL NUMBER 12953631

CRYOSTAT MODEL 7935-7S

PREAMPLIFIER MODEL 2002CSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 10 %
Resolution 1.8 keV (FWHM) at 1.33 MeV
 keV (FWTM) at 1.33 MeV
< 700 keV (FWHM) at 122 keV
 keV (FWTM) at

Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mm
Cryostat description if special

Physical Characteristics

Geometry Closed-end coaxial
Diameter 46 mm Active Volume cc
Length 29.5 mm Well Depth mm
Distance from window 5 mm Well Diameter mm

Electrical Characteristics

Depletion voltage (+)1000 V dc
Recommended bias voltage (+)1500 V dc
Leakage current at recommended bias 0.02 nA
Preamplifier test point voltage at recommended voltage (-)1.59 V dc
Capacitance at recommended bias ~9 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.70	1.75			
FWTM (KeV)	1.32	3.26			
Peak / Compton		38.9:1			
Rel. Efficiency %		10.1			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.5 Liters per Day.

Tested by: Randall Peck Date: December 22, 1995

Approved by: Marjorie Smith Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC1018/SSERIAL NUMBER 12953636CRYOSTAT MODEL 7935-7SPREAMPLIFIER MODEL 2002CSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 10 %Resolution 1.8 keV (FWHM) at 1.33 MeV keV (FWTM) at 1.33 MeV< 700 keV (FWHM) at 122 keV keV (FWTM) at Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mmCryostat description if special

Physical Characteristics

Geometry Closed-end coaxialDiameter 46 mmLength 29 mmDistance from window 5 mmActive Volume ccWell Depth mmWell Diameter mm

Electrical Characteristics

Depletion voltage (+)1200 V dcRecommended bias voltage (+)3000 V dcLeakage current at recommended bias 0.01 nAPreamplifier test point voltage at recommended voltage (-)2.16 V dcCapacitance at recommended bias ~ 8 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.67	1.76			
FWTM (KeV)	1.24	3.35			
Peak / Compton		37.2:1			
Rel. Efficiency %		10.1			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.5 Liters per Day.Tested by: Randall Pech Date: December 22, 1995Approved by: [Signature] Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020SERIAL NUMBER 12953589CRYOSTAT MODEL 7935-7SPREAMPLIFIER MODEL 2101PCSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %Resolution 20 keV (FWHM) at 1.33 MeV keV (FWHM) at 1.33 MeV keV (FWHM) at keV (FWHM) at Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mmCryostat description if special

Physical Characteristics

Geometry Closed-end coaxialDiameter 53 mmLength 44 mmDistance from window 5 mmActive Volume ccWell Depth mmWell Diameter mm

Electrical Characteristics

Depletion voltage (+)3300 V dcRecommended bias voltage (+)3500 V dcLeakage current at recommended bias -- nAReset rate at recommended voltage 0.4 secCapacitance at recommended bias ~19 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.74	1.68			
FWTM (KeV)	1.36	3.10			
Peak / Compton		58.0			
Rel. Efficiency %		23.9			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.Tested by: Randall Peck Date: December 22, 1995Approved by: Margie Laitt Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020SERIAL NUMBER 12953590CRYOSTAT MODEL 7935-7SPREAMPLIFIER MODEL 2101PCSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %Resolution 2.0 keV (FWHM) at 1.33 MeV. keV (FWTM) at 1.33 MeV keV (FWHM) at keV (FWTM) at Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mmCryostat description if special

Physical Characteristics

Geometry Closed-end coaxialDiameter 53 mmLength 45.5 mmDistance from window 5 mmActive Volume ccWell Depth mmWell Diameter mm

Electrical Characteristics

Depletion voltage (+)2400 V dcRecommended bias voltage (+)3000 V dcLeakage current at recommended bias — nAReset rate at recommended voltage 0.6 secCapacitance at recommended bias ~19 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.82	1.77			
FWTM (KeV)	1.53	3.35			
Peak / Compton		54.9			
Rel. Efficiency %		24.0			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.Tested by: Randall Perch Date: December 22, 1995Approved by: [Signature] Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020SERIAL NUMBER 12953592CRYOSTAT MODEL 7935-7SPREAMPLIFIER MODEL 2101PSL

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %Resolution 2.0 keV (FWHM) at 1.33 MeV keV (FWHM) at 1.33 MeV keV (FWHM) at keV (FWHM) at Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mmCryostat description if special

Physical Characteristics

Geometry Closed-end coaxialDiameter 53.5 mmLength 44.5 mmDistance from window 5 mmActive Volume ccWell Depth mmWell Diameter mm

Electrical Characteristics

Depletion voltage (+)3800 V dcRecommended bias voltage (+)4000 V dcLeakage current at recommended bias 0.01 nAReset rate at recommended bias voltage 0.5 secCapacitance at recommended bias ~16 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.69	1.66			
FWTM (KeV)	1.29	3.03			
Peak / Compton		58.2:1			
Rel. Efficiency %		24.2			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.Tested by: John M. Costa Date: December 22, 1995Approved by: M. J. Costa Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020SERIAL NUMBER 12953596CRYOSTAT MODEL 7935-7SPREAMPLIFIER MODEL 2101P

The purchase specification and therefore the warranted performance of this detector are as follows:

(Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %Resolution 2.0 keV (FWHM) at 1.33 MeV keV (FWTM) at 1.33 MeV keV (FWHM) at keV (FWTM) at Peak / Compton 46 :1 Cryostat well diameter mm Cryostat well depth mmCryostat description if special

Physical Characteristics

Geometry Closed-end coaxialDiameter 53 mmLength 43.5 mmDistance from window 5 mmActive Volume ccWell Depth mmWell Diameter mm

Electrical Characteristics

Depletion voltage (+)1500 V dcRecommended bias voltage (+)2500 V dcLeakage current at recommended bias ~ nAReset rate at recommended voltage 0.6 secCapacitance at recommended bias ~18 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.73	1.72			
FWTM (KeV)	1.33	3.18			
Peak / Compton		55.6			
Rel. Efficiency %		22.3			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.Tested by: Randall Pelt Date: December 22, 1995Approved by: M. J. Smith Date: December 22, 1995

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020 SERIAL NUMBER 3963597
 CRYOSTAT MODEL 7935-7S PREAMPLIFIER MODEL 2101PCSL

The purchase specification and therefore the warranted performance of this detector are as follows:
 (Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %
 Resolution 2.0 keV (FWHM) at 1.33 MeV
 keV (FWTM) at 1.33 MeV
 keV (FWHM) at
 keV (FWTM) at

Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mm
 Cryostat description if special Big Mac

Physical Characteristics

Geometry Closed-end coaxial
 Diameter 55 mm Active Volume cc
 Length 42 mm Well Depth mm
 Distance from window 5 mm Well Diameter mm

Electrical Characteristics

Depletion voltage (+)3500 V dc
 Recommended bias voltage (+)4000 V dc
 Reset rate at recommended bias 1.0 sec
 Preamplifier test point voltage at recommended voltage V dc
 Capacitance at recommended bias ~18 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.79	1.75			
FWTM (KeV)	1.46	3.24			
Peak / Compton		53.5:1			
Rel. Efficiency %		23.5			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.

Tested by: Peter G. Jones Date: March 11, 1996

Approved by: Marjorie L. Smith Date: March 11, 1996

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020 SERIAL NUMBER 3963606
CRYOSTAT MODEL 7935-7S PREAMPLIFIER MODEL 2101P

The purchase specification and therefore the warranted performance of this detector are as follows:
(Electric cooling may degrade performance by as much as 10%.)

Active Volume	cc	Relative Efficiency	20	%
Resolution	2.0	keV (FWHM) at 1.33 MeV		
		keV (FWTM) at 1.33 MeV		
		keV (FWHM) at		
		keV (FWTM) at		

Peak / Compton _____:1 Cryostat well diameter _____ mm Cryostat well depth _____ mm
Cryostat description if special _____

Physical Characteristics

Geometry		Closed-end coaxial
Diameter	56	mm
Length	43.5	mm
Distance from window	5	mm
Active Volume		cc
Well Depth		mm
Well Diameter		mm

Electrical Characteristics

Depletion voltage (+)1300 V dc
Recommended bias voltage (+)2500 V dc
Leakage current at recommended bias -- nA
Preamplifier test point voltage at recommended voltage -- V dc
Capacitance at recommended bias ~27 pF
Reset Rate 0.2 sec

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.76	1.90			
FWTM (KeV)	1.42	3.65			
Peak / Compton		51.4:1			
Rel. Efficiency %		25.5			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.5 Liters per Day.

Tested by: Mark Beatty Date: March 11, 1996

Approved by: Munir S. Amiri Date: March 11, 1996

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det Spac_1 10/95 JME

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020 SERIAL NUMBER 3963611
 CRYOSTAT MODEL 7935-7S PREAMPLIFIER MODEL 2101PCSL

The purchase specification and therefore the warranted performance of this detector are as follows:
 (Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %

Resolution 2.0 keV (FWHM) at 1.33 MeV

 keV (FWTM) at 1.33 MeV

 keV (FWHM) at

 keV (FWTM) at

Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mm

Cryostat description if special

Physical Characteristics

Geometry Closed-end coaxial

Diameter 55 mm

Length 43.5 mm

Distance from window 5 mm

Active Volume cc

Well Depth mm

Well Diameter mm

Electrical Characteristics

Depletion voltage (+)2400 V dc

Recommended bias voltage (+)3000 V dc

Reset rate at recommended bias 0.5 sec

Preamplifier test point voltage at recommended voltage V dc

Capacitance at recommended bias ~ 19 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.76	1.83			
FWTM (KeV)	1.41	3.43			
Peak / Compton		52.1			
Rel. Efficiency %		25.2			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.

Tested by: Peter Palmer Date: March 11, 1996

Approved by: Manning Smith Date: March 11, 1996

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JMT

DETECTOR SPECIFICATION AND PERFORMANCE DATA

Specifications

DETECTOR MODEL GC2020 SERIAL NUMBER 3963617
 CRYOSTAT MODEL 7935-7S PREAMPLIFIER MODEL 2101PCSL

The purchase specification and therefore the warranted performance of this detector are as follows:
 (Electric cooling may degrade performance by as much as 10%.)

Active Volume cc Relative Efficiency 20 %

Resolution 2.0 keV (FWHM) at 1.33 MeV

 keV (FWTM) at 1.33 MeV

 keV (FWHM) at

 keV (FWTM) at

Peak / Compton :1 Cryostat well diameter mm Cryostat well depth mm

Cryostat description if special

Physical Characteristics

Geometry Closed-end coaxial
 Diameter 53.5 mm Active Volume cc
 Length 43 mm Well Depth mm
 Distance from window 5 mm Well Diameter mm

Electrical Characteristics

Depletion voltage (+) 5000 V dc
 Recommended bias voltage (+)5000 V dc
 Reset rate at recommended bias 0.4 sec
 Preamplifier test point voltage at recommended voltage V dc
 Capacitance at recommended bias ~19 pF

Resolution and Efficiency

With amplifier time constant of 4 microseconds

Isotope	⁵⁷ Co	⁶⁰ Co			
Energy (KeV)	122	1332			
FWHM (KeV)	0.84	1.73			
FWTM (KeV)	1.57	3.17			
Peak / Compton		56.1:1			
Rel. Efficiency %		23.9			

Cool Down Time 4 hours. Cryostat Liquid Nitrogen Consumption Rate <1.4 Liters per Day.

Tested by: Randall Park Date: March 11, 1996

Approved by: Manning L. Lachett Date: March 11, 1996

800 Research Parkway, Meriden CT. U.S.A. 06450 Tel. 203-238-2351/Fax 203-235-1347

Det_Spec_1 10/95 JHT

APPENDIX D

Energy and FWHM

Data and Curves

HNF-5148, Rev. 2

Energy Calibration Report 12-20-99 11:59:21 AM Page I
 E N B R G Y C A L I B R A T I O N R E P O R T

Identification Number : CALCOUNT
 Description/Title : Calibration Count Counter : GEA - A
 Arrangement : 55-gal. drums Calibration for : SEGE1
 Acquisition Started : 12-20-99 11:31:21 AM
 Live Time 595.4 seconds Real Time 600.0 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 12-20-99 11:59:16 AM by: System Manager
 Energy Calibrate Type: POLY

$$\text{Energy(keV)} = -0.983 + 0.500 \cdot \text{ch} + -1.68\text{E-}07 \cdot \text{ch}^2 + 0.00\text{E+}00 \cdot \text{ch}^3$$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 12-20-99 11:59:16 AM
 by: System Manager

FWHM 0.534 + 0.030*E^{1/2}
 LOW TAIL 0.0E+00 + 0.0E+00*E

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
121.12	0.03	59.54
164.04	0.01	81.00
555.15	0.03	276.40
608.05	0.02	302.84
714.44	0.01	356.01
770.21	0.03	383.85
1326.44	0.03	661.66
2351.34	0.06	1173.24
2670.46	0.06	1332.50

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	1.63	0.06	0.00	0.00
81.00	1.63	0.03	0.00	0.00
276.40	2.11	0.07	0.00	0.00
302.84	2.05	0.04	0.00	0.00
356.01	2.15	0.02	0.00	0.00
383.85	2.21	0.07	0.00	0.00
661.66	2.67	0.07	0.00	0.00
1173.24	3.33	0.11	0.00	0.00
1332.50	3.43	0.11	0.00	0.00

HNF-5148, Rev.2

Energy Calibration Report 12-20-99 11:57:24 AM Page I
 E N E R G Y C A L I B R A T I O N R E P O R T

identification Number : CALCOUNT
 Description/Title : Calibration Count Counter : GEA - A
 Arrangement : 55-gal. drums Calibration for : SEGE2
 Acquisition Started : 12-20-99 11:31:21 AM
 Live Time 592.5 seconds Real Time 600.0 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 12-20-99 11:57:18 AM by: System Manager
 Energy Calibrate Type: POLY

$$\text{Energy(keV)} = -0.757 + 0.500 \cdot \text{ch} + -1.72\text{E-}07 \cdot \text{ch}^2 + 0.00\text{E+}00 \cdot \text{ch}^3$$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 12-20-99 11:57:18 AM
 by: System Manager

$$\text{FWHM} = 0.387 + 0.039 \cdot \text{E}^{1/2}$$

$$\text{LOW TAIL} = 0.0 \cdot \text{E}^{+00} + 0.0\text{E+}00 \cdot \text{ENERGY}$$

CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
120.52	0.02	59.54
163.46	0.01	81.00
554.12	0.02	276.40
607.06	0.01	302.84
713.37	0.01	356.01
769.07	0.02	383.85
1324.89	0.02	661.66
2348.89	0.04	1173.24
2667.67	0.05	1332.50

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	1.64	0.04	0.00	0.00
81.00	1.50	0.02	0.00	0.00
276.40	2.09	0.04	0.00	0.00
302.84	2.11	0.02	0.00	0.00
356.01	2.22	0.01	0.00	0.00
383.85	2.28	0.04	0.00	0.00
661.66	2.79	0.04	0.00	0.00
1173.24	3.80	0.08	0.00	0.00
1332.50	4.16	0.09	0.00	0.00

HNF-5148, Rev. 2

Energy Calibration Report 12-20-99 11:55:41 AM Page I
 E N E R G Y C A L I B R A T I O N R E P O R T

Identification Number : CALCOUNT
 Description/Title : Calibration Count Counter : GEA - A
 Arrangement : 55-gal. drums Calibration for : 8EGE3
 Acquisition Started : 12-20-99 11:31:21 AM
 Live Time 592.6 seconds Real Time 600.0 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 12-20-99 11:55:36 AM by: System Manager
 Energy Calibrate Type: POLY

Energy(keV) = $-0.693 + 0.500 \cdot ch + -1.64E-07 \cdot ch^2 + 0.00E+00 \cdot ch^3$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 12-20-99 11:55:36 AM
 by: System Manager

FWHM = $0.543 + 0.032 \cdot E^{1/2}$
 LOW TAIL = $0.0E+00 + 0.0E+00 \cdot E$

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
120.45	0.02	59.54
163.40	0.01	81.00
554.23	0.02	276.40
607.17	0.01	302.84
713.55	0.01	356.01
769.22	0.02	383.85
1325.28	0.02	661.66
2349.63	0.04	1173.24
2668.63	0.04	1332.50

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	1.72	0.04	0.00	0.00
81.00	1.68	0.02	0.00	0.00
276.40	2.12	0.04	0.00	0.00
302.84	2.18	0.02	0.00	0.00
356.01	2 '25	0.01	0.00	0.00
383.85	2.20	0.04	0.00	0.00
661.66	2.76	0.04	0.00	0.00
1173.24	3.48	0.07	0.00	0.00
1332.50	3.73	0.08	0.00	0.00

HNF-5148, Rev. 2

Energy Calibration Report 12-20-99 11:53:53 AM Page I
 E N E R G Y C A L I B R A T I O N R E P O R T

identification Number CALCOUNT
 Description/Title Calibration Count Counter GEA - A
 Arrangement 55-gal. drums Calibration for SEGE4
 Acquisition Started 12-20-99 11:31:21 AM Live Time
 595.0 seconds Real Time 600.0 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 12-20-99 11:53:48 AM by: System Manager
 Energy Calibrate Type: POLY

$$\text{Energy(keV)} = -1.687 + 0.501 \cdot \text{ch} + -1.51\text{E-}07 \cdot \text{ch}^2 + 0.00\text{E+}00 \cdot \text{ch}^3$$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 12-20-99 11:53:48 AM
 by: System Manager 9*EA

$$\begin{aligned} \text{FWHM} &= 0.529 + 0.02 \cdot \text{E}^{1/2} \\ \text{LOW TAIL} &= 0.0\text{E+}00 + 0.0\text{E+}00 \cdot \text{E} \end{aligned}$$

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
122.31	0.02	59.54
165.21	0.01	81.00
555.68	0.03	276.40
608.49	0.02	302.84
714.77	0.01	356.01
770.34	0.03	383.85
1325.81	0.03	661.66
2348.89	0.05	1173.24
2667.71	0.05	1332.50

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	1.61	0.05	0.00	0.00
81.00	1.59	0.02	1.00	0.00
276.40	2.00	0.06	0.00	0.00
302.84	2.05	0.03	0.00	0.00
356.01	2.13	0.02	0.00	0.00
383.85	2.16	0.06	0.00	0.00
661.66	2.52	0.05	0.00	0.00
1173.24	3.21	0.09	0.00	0.00
1332.50	3.33	0.09	0.00	0.00

HNF-5148, Rev. 2

Energy Calibration Report 12-20-99 12:01:44 PM Page I
E N E R G Y C A L I B R A T I O N R E P O R T

Identification Number : CALCOUNT
Description/Title : Calibration Count Counter : GEA - A
Arrangement : 55-gal. drums Calibration for : SEGES
Acquisition Started : 12-20-99 11:31:21 AM
Live Time 595.0 seconds Real Time 600.0 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 12-20-99 12:01:35 PM by: System Manager
Energy Calibrate Type: POLY

$$\text{Energy(keV)} = -1.691 + 0.501 \cdot \text{ch} + -1.54\text{E-}07 \cdot \text{ch}^2 + 0.00\text{E+}00 \cdot \text{ch}^3$$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 12-20-99 12:01:35 PM
by: System Manager

$$\text{FWHM} = 0.605 + 0.033 \cdot \text{E}^{1/2}$$

$$\text{LOW TAIL} = 0.0\text{E+}00 + 0.0\text{E+}00 \cdot \text{E}$$

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
122.31	0.01	59.54
165.22	0.01	81.00
555.64	0.01	276.40
608.51	0.01	302.84
714.76	0.00	356.01
770.38	0.01	383.85
1325.82	0.01	661.66
2348.98	0.02	1173.24
2667.61	0.02	1332.50

SHAPE CALI13RATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	1.86	0.03	0.00	0.00
81.00	1.80	0.01	0.00	0.00
276.40	2.35	0.03	0.00	0.00
302.84	2.36	0.02	0.00	0.00
356.01	2.41	0.01	0.00	0.00
383.85	2.40	0.03	0.00	0.00
661.66	2.85	0.02	0.00	0.00
1173.24	3.63	0.04	0.00	0.00
1332.50	3.81	0.04	0.00	0.00

Energy Calibration Report

1-31-00 10:47:11 AM

Page I

E N E R G Y C A L I B R A T I O N R E P O R T

Identification Number : CA-LCOUNT
 Description/Title : Calibration Count Counter : GEA - A
 Arrangement : 55-gal. drums.
 Calibration for : LEGES
 Acquisition Started : 1-31-00 10:33:40 AM
 Live Time 299.8 seconds Real Time 306.4 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 1-31-00 10:46:44 AM
 by: System Manager
 Energy Calibrate Type: POLY

Energy(keV) = $0.145 + 0.077 \cdot \text{ch} + 0.00\text{E}+00 \cdot \text{ch}^2 + 0.00\text{E}+00 \cdot \text{ch}^3$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 1-31-00 10:46:44 AM
 by: System Manager

FWHM = $0.306 + 0.032 \cdot E^{1/2}$
 LOW TAIL = $0.0\text{E}+00 + 0.0\text{E}+00 \cdot E$

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
774.51	0.46	59.54
1054.24	0.11	81.00
3601.60	0.12	276.40
3946.74	0.07	302.84

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	5.61	1.05	0.00	0.00
81.00	7.80	0.23	0.00	0.00
276.40	11.38	0.25	0.00	0.00
302.84	11.13	0.14	0.00	0.00

Energy Calibration Report

1-31-00 10:53:25 AM

Page I

E N E R G Y C A L I B R A T I O N R E P O R T

Identification Number :CALCOUNT
 Description/Title :Calibration Count Counter :GEA - A
 Arrangement :55-gal. drums Calibration for :LEGE6
 Adquisition Started 1-31-00 10:33:40 AM
 Live Time 300.6 seconds Real Time 306.8 seconds

ENERGY CALIBRATION COEFFICIENTS

Energy Calibrate Performed on: 1-31-00 10:53:13 AM
 by: System Manager
 Energy Calibrate Type: POLY

Energy(keV) = $0.118 + 0.076 \cdot \text{ch} + 0.00\text{E}+00 \cdot \text{ch}^2 + 0.00\text{E}+00 \cdot \text{ch}^3$

SHAPE CALIBRATION COEFFICIENTS

Shape Calibrate Performed on: 1-31-00 10:53:13 AM
 by: System Manager

FWHM = $0.241 + 0.040 \cdot E^{1/2}$

LOW TAIL = $0.0\text{E}+00 + 0.0\text{E}+00 \cdot E$

ENERGY CALIBRATION RESULTS TABLE

Centroid Channel	Centroid error	Energy (keV)
778.08	0.67	59.54
1058.65	0.11	81.00
3616.11	0.13	276.40
3962.30	0.08	302.84

SHAPE CALIBRATION RESULTS TABLE

Energy (keV)	FWHM channels	FWHM error	TAIL channels	TAIL error
59.54	4.90	1.38	0.00	0.00
81.00	8.00	0.24	0.00	0.00
276.40	11.72	0.25	0.00	0.00
302.84	12.40	0.15	0.00	0.00

APPENDIX E

SEGe Efficiency

Data and Curves

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Foam Drum, density = 0.01346 gms/cc, No Shield (collimator). Spectrum Seq # 3988

Datasource: C:\WAS\TEMP\EFF642.TEM

$\ln(\text{Eff}) = -1.374\text{e}+02 + 8.105\text{e}+01 \cdot \ln(E) - 6.551\text{e}-02 \cdot \ln(E)^4$

$-1.824\text{e}+01 \cdot \ln(E)^2 + 1.793\text{e}+00 \cdot \ln(E)^3$

Sum Segments.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Datasource: C:\WAS\TEMP\EFF642.TEM
 $\ln(\text{Eff}) = -5.168e+02 + 3.223e+02 \ln(E) - 7.662e+01 \ln(E)^2 + 8.080e+00 \ln(E)^3 - 3.190e-01 \ln(E)^4$

Seg 8, SEGe 2, Drum position 2.

Datasource: C:\WAS\TEMP\EFF642.TEM
 $\ln(\text{Eff}) = -1.893e+01 + 3.569e+00 \ln(E) - 3.903e-01 \ln(E)^2 + 1.070e-02 \ln(E)^3$

Seg 9, SEGe 2, Drum position 3.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Foam Drum, density = 0.01346 gms/cc, Shield (collimator). Spectrum Seq # 3989

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Homosote Drum, density = 0.43029 gms/cc, No Shield (collimator). Spectrum Seq # 3994

Datasource: C:\WAS\TEMP\EFF478.TEM

$\ln(\text{Eff}) = -3.182e+01 + 9.945e+00 \ln(E) - 1.161e+00 \ln(E)^2 + 3.601e-02 \ln(E)^3$

Sum Segments

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Datasource: C:\WAS\TEMP\EFF478.TEM
 $\ln(\text{Eff}) = -5.468\text{e}+01 + 1.913\text{e}+01 \cdot \ln(E) - 2.751\text{e}+00 \cdot \ln(E)^2 + 1.314\text{e}-01 \cdot \ln(E)^3$

Seg 2, SEGe 4, Drum position 2,

Datasource: C:\WAS\TEMP\EFF478.TEM
 $\ln(\text{Eff}) = -1.110\text{e}+03 + 6.916\text{e}+02 \cdot \ln(E) - 1.626\text{e}+02 \cdot \ln(E)^2 + 1.695\text{e}+01 \cdot \ln(E)^3 - 6.613\text{e}-01 \cdot \ln(E)^4$

Seg 3, SEGe 4, Drum position 3

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Homosote Drum, density = 0.43029 gms/cc, Shield (collimator). Spectrum Seq # 3995

Datasource: C:\WAS\TEMP\EFF478.TEM

$\ln(\text{Eff}) = -1.259\text{e}+01 + 1.391\text{e}+00 \cdot \ln(E) - 1.002\text{e}-01 \cdot \ln(E)^2$

Sum Segments.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Datasource: C:\WAS\TEMP\EFF039.TEM
 $\ln(\text{Eff}) = -1.377\text{e}+02 + 7.760\text{e}+01 \cdot \ln(E) - 1.720\text{e}+01 \cdot \ln(E)^2 + 1.670\text{e}+00 \cdot \ln(E)^3 - 8.032\text{e}-02 \cdot \ln(E)^4$

Seg 4, SEGe 3, Drum position 1.

Datasource: C:\WAS\TEMP\EFF039.TEM
 $\ln(\text{Eff}) = -1.854\text{e}+02 + 1.133\text{e}+02 \cdot \ln(E) - 2.692\text{e}+01 \cdot \ln(E)^2 + 2.819\text{e}+00 \cdot \ln(E)^3 - 1.103\text{e}-01 \cdot \ln(E)^4$

Seg 5, SEGe 3, Drum position 2.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Datasource: C:\WAS\TEMP\EFF039.TEM
 $\ln(\text{Eff}) = 1.889\text{e}+02 - 2.096\text{e}+02 \ln(E) + 8.263\text{e}+01 \ln(E)^2 - 1.549\text{e}+01 \ln(E)^3$
 $+ 1.400\text{e}+00 \ln(E)^4 - 4.926\text{e}-02 \ln(E)^5$

Seg 8, SEGe 2, Drum position 2.

Datasource: C:\WAS\TEMP\EFF039.TEM
 $\ln(\text{Eff}) = -1.561\text{e}+02 + 9.162\text{e}+01 \ln(E) - 2.111\text{e}+01 \ln(E)^2 + 2.145\text{e}+00 \ln(E)^3$
 $- 8.160\text{e}-02 \ln(E)^4$

Seg 9, SEGe 2, Drum position 3.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3992

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 1, Drum position 3.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Particle Board Drum, density = 0.65625 gms/cc, No Shield (collimator). Spectrum Seq # 3993

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2

Sand Drum, density = 1.5601 gms/cc, No Shield (collimator). Spectrum Seq # 3991

Datasource: C:\WAS\TEMP\EFF039.TEM

$$\ln(\text{Eff}) = -4.273e+01 + 1.457e+01 \cdot \ln(E) - 1.888e+00 \cdot \ln(E)^2 + 7.481e-02 \cdot \ln(E)^3$$

Sum Segments

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Seg 2, SEGe 4, Drum position 2.

Seg 3, SEGe 4, Drum position 3.

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Seg 4, SEGe 3, Drum position 1.

Seg 5, SEGe 3, Drum position 2.

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Seg 6, SEGe 3, Drum position 3.

Seg 7, SEGe 2, Drum position 1.

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Seg 8, SEGe 2, Drum position 2.

Seg 9, SEGe 2, Drum position 3.

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Seg 10, SEGe 1, Drum position 1.

Seg 11, SEGe 1, Drum position 2.

Sand Drum, density = 1.5601 gms/cc Shield (collimator). Spectrum Seq # 3990

Summed Spectra, low density, 0.0135, no shield.

Datasource: C:\WAS\TEMP\EFF478.TEM
 $\ln(\text{Eff}) = -1.397\text{e}+02 + 8.105\text{e}+01 \cdot \ln(E) - 1.824\text{e}+01 \cdot \ln(E)^2 + 1.793\text{e}+00 \cdot \ln(E)^3$
 $-6.551\text{e}-02 \cdot \ln(E)^4$

Summed Spectra, 2nd density, 0.430, no shield.

Datasource: C:\WAS\TEMP\EFF478.TEM
 $\ln(\text{Eff}) = -2.260\text{e}+02 + 1.417\text{e}+02 \cdot \ln(E) - 3.423\text{e}+01 \cdot \ln(E)^2 + 3.644\text{e}+00 \cdot \ln(E)^3$
 $-1.449\text{e}-01 \cdot \ln(E)^4$

Summed Spectra, 3rd density, 0.656, no shield.

Summed Spectra, heavy density, 1.56, no shield.

APPENDIX F

Spreadsheets

Spreadsheet for Figure 2

Activity and Count response for 1-2000 cal

Daily Source Drum

COUNTS									
Seg #	Seg	59 k'eV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11-Top	S12	4542	4583	4537	4331	4617	4522.0	111.64	2.47
10	S13	6777	6588	6669	9834	6678	7309.2	1413.00	19.33
9	S21	7964	8149	8377	8470	8370	8266.0	205.94	2.49
8	S22	7992	8244	7991	8124	8320	8134.2	147.82	1.82
7	S23	8686	8591	8359	8416	8572	8524.8	134.09	1.57
6	S31	7949	8185	8197	8296	8269	8179.2	136.98	1.67
5	S32	7504	7763	7784	7625	7416	7618.4	160.02	2.10
4	S33	7134	7406	7283	6855	7064	7148.4	210.88	2.95
3	S41	5330	5732	5591	5716	5701	5614.0	168.14	2.99
2	S42	3138	3322	3358	3273	3209	3260.0	88.18	2.70
A04		69025	69620	70322	70768	70642	70075.4	736.93	1.05
Sum		67016	68563	68146	70940	68216	68576.2		
% Diff		2.91	1.52	3.09	-0.24	3.43	2.14		

ACTIVITY									
Seg #	Seg	59 k'eV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11	S12	4.034	4.076	4.036	3.853	4.106	4.021	0.10	2.46
10	S13	3.765	3.660	3.705	3.797	3.711	3.728	0.05	1.44
9	S21	3.634	3.717	3.821	3.864	3.819	3.771	0.09	2.49
8	S22	3.831	3.952	3.831	3.895	3.988	3.899	0.07	1.82
7	S23	3.564	3.525	3.430	3.453	3.517	3.498	0.06	1.58
6	S31	3.512	3.616	3.621	3.665	3.653	3.613	0.06	1.67
5	S32	3.761	3.891	3.901	3.822	3.717	3.818	0.08	2.10
4	S33	3.794	3.938	3.873	3.645	3.756	3.801	0.11	2.95
3	S41	3.275	3.522	3.436	3.512	3.503	3.449	0.10	3.00
2	S42	3.404	3.603	3.642	3.550	3.481	3.536	0.10	2.70
A04		36.707	37.024	37.398	37.635	37.568	37.266	0.39	1.05
Sum		36.573	37.501	37.297	37.055	37.251	37.135		
% Diff		0.37	-1.29	0.27	1.54	0.84	0.35		

Activity and Count response for 1-2000 cal

Daily Source Drum

COUNTS									
Seg #	Seg	81 keV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11-Top	S12	12013	11293	12379	12377	11693	11951.0	465.84	3.90
10	S13	18923	18848	18870	18727	19251	18923.8	196.49	1.04
9	S21	19042	22199	18865	20424	18564	19818.8	1509.91	7.62
8	S22	19992	20528	19281	21952	22377	20826.0	1308.13	6.28
7	S23	18957	20505	20621	20572	19577	20046.4	745.63	3.72
6	S31	19310	19039	19336	18760	19429	19174.8	273.58	1.43
5	S32	18497	18676	18039	18786	18423	18484.2	287.22	1.55
4	S33	16774	17488	17278	17679	17312	17306.2	337.67	1.95
3	S41	13183	11923	15999	16188	11539	13766.4	2210.68	16.06
2	S42	7167	9236	9220	9042	7973	8527.6	922.59	10.82
A04		149800	115291	150812	151217	151793	143782.6	15943.88	11.09
Sum		163858	169735	169888	174507	166138	168825.2		
% Diff		-9.38	-47.22	-12.65	-15.40	-9.45	-17.42		

ACTIVITY									
Seg #	Seg	81 keV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11	S12	4.483	4.214	4.620	4.619	4.363	4.460	0.17	3.90
10	S13	3.579	3.565	3.569	3.542	3.641	3.579	0.04	1.04
9	S21	2.899	3.380	2.872	3.109	2.827	3.017	0.23	7.61
8	S22	3.395	3.486	3.274	3.728	3.800	3.537	0.22	6.28
7	S23	3.152	3.410	3.429	3.412	3.256	3.332	0.12	3.68
6	S31	3.476	3.427	3.480	3.376	3.497	3.451	0.05	1.43
5	S32	3.522	3.556	3.435	3.577	3.508	3.520	0.05	1.55
4	S33	3.196	3.332	3.292	3.368	3.298	3.297	0.06	1.95
3	S41	2.996	2.710	3.637	3.680	2.623	3.129	0.50	16.06
2	S42	3.019	3.891	3.884	3.809	3.359	3.592	0.39	10.82
A04		34.010	34.349	34.240	34.332	34.463	34.279	0.17	0.50
Sum		33.717	34.970	35.492	36.220	34.171	34.914		
% Diff		0.86	-1.81	-3.66	-5.50	0.85	-1.85		

Activity and Count response for 1-2000 cal

Daily Source Drum

COUNTS

Seg #	Seg	356 keV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11-Top	S12	25224	25856	25486	25891	25730	25637.4	280.4	1.09
10	S13	38324	39185	38780	38893	38746	38785.6	310.5	0.80
9	S21	43236	42259	43578	43350	43331	43150.8	514.1	1.19
8	S22	43600	43748	43871	44066	44089	43874.8	208.6	0.48
7	S23	44396	43356	44099	44096	43949	43979.2	384.4	0.87
6	S31	46295	46896	46348	46799	47071	46681.8	343.6	0.74
5	S32	46082	45829	45734	46506	45967	46023.6	300.4	0.65
4	S33	43308	43777	43397	43272	43427	43436.2	200.7	0.46
3	S41	32772	33185	32852	32872	32803	32896.8	165.9	0.50
2	S42	19447	19879	19622	19376	19497	19564.2	197.6	1.01
A04		382713	385298	383730	385822	386017	384716	1435.2	0.37
Sum		382684	383970	383767	385121	384610	384030.4		
% Diff		0.01	0.34	-0.01	0.18	0.36	0.18		

ACTIVITY

Seg #	Seg	356 keV		Sequence number			AVG	Dev	% err
		4106	4107	4108	4109	4110			
11	S12	3.080	3.157	3.113	3.162	3.142	3.131	0.034	1.09
10	S13	3.045	3.113	3.081	3.090	3.078	3.082	0.025	0.80
9	S21	3.139	3.141	3.164	3.148	3.146	3.148	0.010	0.31
8	S22	3.104	3.114	3.124	3.137	3.139	3.124	0.015	0.48
7	S23	3.161	3.087	3.140	3.140	3.130	3.132	0.027	0.87
6	S31	3.118	3.159	3.122	3.152	3.171	3.144	0.023	0.73
5	S32	3.186	3.168	3.162	3.215	3.178	3.182	0.021	0.65
4	S33	3.099	3.133	3.106	3.097	3.108	3.108	0.014	0.46
3	S41	3.149	3.189	3.157	3.159	3.152	3.161	0.016	0.51
2	S42	3.206	3.277	3.235	3.193	3.214	3.225	0.033	1.02
A04		31.944	32.160	32.030	32.204	32.221	32.112	0.120	0.37
Sum		31.288	31.540	31.404	31.492	31.458	31.436		
% Diff		2.05	1.93	1.95	2.21	2.37	2.10		

Activity and Count response for 1-2000 cal

Daily Source Drum

COUNTS

Seg #	Seg	661 keV	Sequence number					AVG	Dev	% err
		4106	4107	4108	4109	4110				
11-Top	S12	4062	4026	3939	3922	3983	3986.4	58.50	1.47	
10	S13	5767	5929	6034	6082	5894	5941.2	123.65	2.08	
9	S21	6856	6646	6685	6695	6719	6720.2	80.35	1.20	
8	S22	6873	6657	6908	6733	6641	6762.4	122.62	1.81	
7	S23	6936	6945	6769	6850	6845	6869.0	72.80	1.06	
6	S31	7308	7204	7178	7225	7250	7233.0	49.61	0.69	
5	S32	7323	7143	7162	7121	7291	7208.0	92.23	1.28	
4	S33	6624	6741	6815	6914	6799	6778.6	106.52	1.57	
3	S41	4957	5035	4980	5032	5024	5005.6	35.06	0.70	
2	S42	2991	2943	3046	3061	3021	3012.4	47.01	1.56	
A04		59842	59656	59766	59750	59553	59713.4	111.43	0.19	
Sum		59697	59269	59516	59635	59467	59516.8			
% Diff		0.24	0.65	0.42	0.19	0.14	0.33			

ACTIVITY

Seg #	Seg	661 keV	Sequence number					AVG	Dev	% err
		4106	4107	4108	4109	4110				
11	S12	0.589	0.584	0.572	0.562	0.578	0.577	0.011	1.89	
10	S13	0.531	0.546	0.556	0.560	0.543	0.547	0.011	2.08	
9	S21	0.575	0.557	0.560	0.561	0.563	0.563	0.007	1.20	
8	S22	0.589	0.570	0.592	0.577	0.569	0.579	0.010	1.81	
7	S23	0.575	0.575	0.561	0.567	0.567	0.569	0.006	1.06	
6	S31	0.608	0.599	0.597	0.601	0.603	0.601	0.004	0.69	
5	S32	0.597	0.582	0.584	0.580	0.594	0.588	0.008	1.28	
4	S33	0.544	0.553	0.560	0.568	0.558	0.557	0.009	1.57	
3	S41	0.557	0.566	0.560	0.566	0.565	0.562	0.004	0.70	
2	S42	0.562	0.553	0.573	0.575	0.568	0.566	0.009	1.56	
A04		5.948	5.929	5.940	5.939	5.919	5.935	0.011	0.19	
Sum		5.726	5.687	5.713	5.717	5.708	5.710			
% Diff		3.72	4.09	3.83	3.73	3.57	3.79			

Activity and Count response for 1-2000 cal

Daily Source Drum

COUNTS									
Seg #	Seg	1332.5 keV	Sequence number				AVG	Dev	% err
		4106	4107	4108	4109	4110			
11-Top	S12	1966	1903	1921	1952	1894	1927.2	31.01	1.61
10	S13	2780	2894	2854	2823	2879	2846.0	45.67	1.60
9	S21	3219	3150	3196	3222	3271	3211.6	43.96	1.37
8	S22	3244	3263	3297	3346	3177	3265.4	62.80	1.92
7	S23	3388	3339	3362	3343	3360	3358.4	19.40	0.58
6	S31	3493	3448	3520	3544	3428	3486.6	48.41	1.39
5	S32	3438	3356	3443	3472	3490	3439.8	51.46	1.50
4	S33	3241	3263	3295	3197	3147	3228.6	57.89	1.79
3	S41	2411	2310	2364	2368	2401	2370.8	39.62	1.67
2	S42	1426	1449	1439	1359	1449	1424.4	37.76	2.65
A04		28645	28393	28746	28562	28554	28580.0	130.14	0.46
Sum		28606	28375	28691	28626	28496	28558.8		
% Diff		0.14	0.06	0.19	-0.22	0.20	0.07		

ACTIVITY									
Seg #	Seg	1332.5 keV	Sequence number				AVG	Dev	% err
		4106	4107	4108	4109	4110			
11	S12	0.4158	0.4024	0.4063	0.4130	0.4007	0.4076	0.0066	1.61
10	S13	0.3647	0.3797	0.3745	0.3703	0.3777	0.3734	0.0060	1.61
9	S21	0.3964	0.3879	0.3936	0.3968	0.4027	0.3955	0.0054	1.37
8	S22	0.3916	0.3940	0.3980	0.4040	0.3835	0.3942	0.0076	1.93
7	S23	0.4115	0.4055	0.4083	0.4060	0.4081	0.4079	0.0024	0.58
6	S31	0.3845	0.3795	0.3874	0.3900	0.3772	0.3837	0.0053	1.39
5	S32	0.4103	0.4006	0.4109	0.4143	0.4165	0.4105	0.0061	1.49
4	S33	0.3899	0.3926	0.3963	0.3845	0.3786	0.3884	0.0070	1.79
3	S41	0.4070	0.3901	0.3993	0.3998	0.4055	0.4003	0.0067	1.66
2	S42	0.3665	0.3723	0.3699	0.3494	0.3724	0.3661	0.0096	2.64
A04		4.0484	4.0128	4.0627	4.0366	4.0356	4.0392	0.0184	0.45
Sum		3.9382	3.9046	3.9445	3.9281	3.9231	3.9277		
% Diff		2.72	2.70	2.91	2.69	2.79	2.76		

Calculation of microCuries per segment detector:

Seg#	Seg	59keV		81keV		356keV		661keV		1332.5keV	
		AVG	%err	AVG	%err	AVG	%err	AVG	%err	AVG	%err
11	S12	402	2.46	446	3.90	313	1.61	0.577	1.89	0.408	1.61
10	S13	373	1.44	3.58	1.04	3.08	1.61	0.547	2.08	0.373	1.61
9	S21	377	2.49	3.02	7.61	3.15	1.37	0.563	1.20	0.395	1.37
8	S22	390	1.82	3.54	6.28	3.12	1.93	0.579	1.81	0.394	1.93
7	S23	350	1.58	3.33	3.68	3.13	0.58	0.559	1.06	0.408	0.58
6	S31	361	1.67	3.45	1.43	3.14	1.39	0.601	0.69	0.384	1.39
5	S32	382	2.10	3.52	1.55	3.18	1.49	0.588	1.28	0.411	1.49
4	S33	380	2.95	3.30	1.95	3.11	1.79	0.557	1.57	0.388	1.79
3	S41	345	3.00	3.13	16.06	3.16	1.66	0.562	0.70	0.400	1.66
2	S42	354	2.70	3.59	10.82	3.22	2.64	0.566	1.56	0.366	2.64

Comparison of Segment Summation to the Summed Spectra results

A04		37.27	1.06	34.28	0.50	32.11	0.45	5.93	0.19	4.04	0.45
Sum	%Diff	37.14	0.35	34.91	-1.85	31.44	2.10	5.71	3.79	3.93	2.76

Ratio to NSI source Activity

ScAd		3597	2.77	3162	2.91	3162	2.91	557	3.42	394	3.05
Ratio A04/Sd		1.036		1.084		1.016		1.066		1.025	
Sum/Sd		1.032		1.104		0.994		1.025		0.997	

No Detector Shield (Coll=0)							
Density	Detector	Drum Position	Coefficients				
			a	b	c	d	e
0.01346	SEGe-4	2	-1.348E+02	7.694E+01	-1.713E+01	1.644E+00	-6.016E-02
	SEGe-4	3	-2.314E+02	1.462E+02	-3.530E+00	3.751E+00	-1.488E-01
	SEGe-3	1	-2.567E+02	1.650E+02	-4.040E+01	4.356E+00	-1.753E-01
	SEGe-3	2	-2.085E+02	1.295E+02	-3.076E+01	3.210E+00	-1.249E-01
	SEGe-3	3	-1.841E+02	1.139E+02	-2.712E+01	2.841E+00	-1.113E-01
	SEGe-2	1	-1.415E+02	8.126E+01	-1.794E+01	1.714E+00	-6.022E-03
	SEGe-2	2	-1.787E+02	1.086E+02	-2.545E+01	2.621E+00	-1.009E-01
	SEGe-2	3	-2.074E+02	1.293E+02	-3.095E+01	3.264E+00	-1.287E-01
	SEGe-1	1	-2.103E+02	1.317E+02	-3.170E+01	3.360E+00	-1.331E-01
	SEGe-1	2	-1.854E+02	1.134E+02	-2.679E+01	2.783E+00	-1.079E-01
	Summed		-1.374E+02	8.105E+01	-1.824E+01	1.793E+00	-6.551E-02
0.43029	SEGe-4	2	-7.526E+01	3.279E+01	-5.305E+00	2.802E-01	
	SEGe-4	3	-7.233E+01	3.153E+01	-5.089E+00	2.677E-01	
	SEGe-3	1	-9.277E+01	4.810E+01	-9.974E+00	8.926E-01	-2.929E-02
	SEGe-3	2	-7.058E+01	3.039E+01	-4.849E+00	2.520E-01	
	SEGe-3	3	-6.257E+01	2.616E+01	-4.123E+00	2.113E-01	
	SEGe-2	1	-2.659E+02	1.685E+02	-4.084E+01	4.359E+00	-1.734E-01
	SEGe-2	2	-7.354E+01	3.199E+01	-5.135E+00	2.690E-01	
	SEGe-2	3	-1.167E+02	6.400E+01	-1.390E+01	1.320E+00	-4.664E-02
	SEGe-1	1	-1.253E+02	7.022E+01	-1.558E+01	1.519E+00	-5.546E-02
	SEGe-1	2	-6.935E+01	2.906E+01	-4.584E+00	2.354E-01	
	Summed		-3.182E+01	9.945E+01	-1.161E+00	3.601E-01	
0.65625	SEGe-4	2	-1.389E+02	7.910E+01	-1.773E+01	1.735E+00	6.289E-02
	SEGe-4	3	-2.480E+02	1.562E+02	-3.779E+01	4.029E+00	-1.602E-01
	SEGe-3	1	-1.377E+02	7.760E+01	-1.720E+01	1.670E+00	-6.032E-02
	SEGe-3	2	-1.854E+02	1.133E+02	-2.692E+01	2.819E+00	1.103E-01
	SEGe-3	3	-1.925E+02	1.179E+02	-2.801E+01	2.933E+00	-1.147E-01
	SEGe-2	1	-7.015E+01	2.983E+01	-4.745E+00	2.469E-01	
	SEGe-2	2	1.889E+02	2.096E+02	9.263E+01	-1.549E+01	1.400E+00
	SEGe-2	3	-1.561E+02	9.162E+01	-2.111E+01	2.145E+00	-8.160E-02
	SEGe-1	1	-1.474E+02	8.440E+01	-1.901E+01	1.852E+00	-6.963E-02
	SEGe-1	2	-2.128E+02	1.300E+02	-3.080E+01	3.217E+00	-1.255E-01
	Summed		-1.686E+02	1.021E+02	-2.382E+01	2.446E+00	-9.376E-02
1.56010	SEGe-4	2	-7.228E+01	2.937E+01	-4.532E+00	2.287E-01	
	SEGe-4	3	-2.179E+02	1.346E+02	-3.229E+01	3.523E+00	-1.357E-01
	SEGe-3	1	-1.574E+02	8.997E+01	-2.015E+01	1.975E+00	-7.186E-02
	SEGe-3	2	-2.116E+02	1.290E+02	-3.059E+01	3.202E+00	-1.253E-01
	SEGe-3	3	-1.738E+02	1.021E+02	-2.349E+01	2.378E+00	-8.979E-02
	SEGe-2	1	-6.852E+01	2.813E+01	-4.383E+00	2.234E-01	
	SEGe-2	2	-6.891E+01	2.805E+01	-4.336E+00	2.193E-01	
	SEGe-2	3	-2.239E+02	1.364E+02	-3.220E+01	3.354E+00	-1.305E-01
	SEGe-1	1	-7.297E+01	3.019E+01	-4.708E+00	2.398E-01	
	SEGe-1	2	-9.201E+01	4.011E+01	-6.435E+00	3.385E-01	
	Summed		-4.273E+01	1.457E+01	-1.888E+00	7.481E-02	
0.01346	Summed		-1.397E+02	8.105E+01	-1.824E+01	1.793E+00	-6.551E-02
0.43029	Summed		-2.260E+02	1.417E+02	-3.423E+01	3.644E+00	-1.449E-01
0.65625	Summed		-1.709E+02	1.021E+02	-2.382E+01	2.446E+00	-9.367E-02
1.56010	Summed		-2.595E+02	1.628E+02	-3.937E+01	4.197E+00	-1.670E-01

Detector Shield (Coll=1)

Density	Detector	Drum Position	Coefficients				
			a	b	c	d	e
0.01346	SEGe-4	2	-6.420E+01	2.566E+01	-4.001E+00	2.061E-01	
	SEGe-4	3	-9.956E+01	4.232E+01	-6.565E+00	3.365E-01	
	SEGe-3	1	-5.473E+01	2.092E+01	-3.175E+00	1.589E-01	
	SEGe-3	2	-8.927E+01	3.665E+01	-5.556E+00	2.786E-01	
	SEGe-3	3	-2.638E+02	-1.657E+02	-4.030E+01	4.346E+00	-1.754E-01
	SEGe-2	1	-5.764E+01	2.261E+01	-3.490E+00	1.779E-01	
	SEGe-2	2	-5.166E+02	3.223E+02	-7.662E+01	8.080E+00	-3.190E-01
	SEGe-2	3	-1.893E+01	3.569E+00	-3.903E-01	1.070E-02	
	SEGe-1	1	-4.623E+01	1.693E+01	-2.573E+00	1.284E-01	
	SEGe-1	2	-5.738E+01	3.434E+01	-9.082E+00	1.046E+00	-4.452E-02
	Summed		-4.043E+01	1.530E+01	-2.311E+00	1.147E-01	
0.43029	SEGe-4	2	-5.468E+01	1.913E+01	-2.751E+00	1.314E-01	
	SEGe-4	3	-1.110E+03	6.916E+02	-1.626E+02	1.695E+01	-6.613E-01
	SEGe-3	1	-1.527E+01	1.589E+00	-1.182E-01		
	SEGe-3	2	-3.973E+02	2.539E+02	-6.238E+01	6.797E+00	-2.770E-01
	SEGe-3	3	-1.234E+02	7.655E+02	-1.792E+02	1.861E+01	-7.232E-01
	SEGe-2	1	-1.959E+01	2.939E+00	-2.223E-01		
	SEGe-2	2	-1.993E+01	3.047E+00	-2.318E-01		
	SEGe-2	3	-1.602E+01	1.747E+00	-1.250E-01		
	SEGe-1	1	-7.623E+02	4.539E+02	-1.025E+02	1.026E+01	-3.843E-01
	SEGe-1	2	-1.740E+01	1.851E+00	-1.279E-01		
	Summed		-1.259E+01	1.391E+00	-1.002E-01		
0.65625	SEGe-4	2	-2.284E+01	3.564E+00	-2.599E-01		
	SEGe-4	3	-1.439E+01	1.249E+00	-9.517E-02		
	SEGe-3	1	-4.610E+02	2.824E+02	-6.626E+01	6.902E+00	-2.690E-01
	SEGe-3	2	-2.031E+01	3.018E+00	-2.244E-01		
	SEGe-3	3	-1.263E+03	7.705E+02	-1.775E+02	1.814E+01	-6.944E-01
	SEGe-2	1	-9.370E+01	5.363E+01	-1.302E+01	1.413E+00	-5.762E-02
	SEGe-2	2	-1.920E+01	2.662E+00	-1.949E-01		
	SEGe-2	3	-1.985E+01	2.861E+00	-2.108E-01		
	SEGe-1	1	-2.343E+03	1.450E+03	-3.374E+02	3.481E+01	-1.344E+00
	SEGe-1	2	-1.158E+03	7.239E+02	-1.709E+02	1.789E+01	-6.999E-01
	Summed		-1.233E+02	6.828E+01	-1.522E+01	1.510E+00	-5.626E-02
1.56010	SEGe-4	2	-2.284E+01	3.564E+00	-2.599E-01		
	SEGe-4	3	-1.439E+01	1.249E+00	-9.517E-02		
	SEGe-3	1	-4.610E+02	2.824E+02	-6.626E+01	6.902E+00	-2.690E-01
	SEGe-3	2	-2.031E+01	3.018E+00	-2.244E-01		
	SEGe-3	3	-1.263E+01	7.705E+02	-1.775E+02	1.814E+01	-6.944E-01
	SEGe-2	1	-9.370E+01	5.363E+01	-1.302E+01	1.413E+00	-5.762E-02
	SEGe-2	2	-1.920E+01	2.660E+00	-1.949E-01		
	SEGe-2	3	-1.985E+01	2.861E+00	-2.108E-01		
	SEGe-1	1	-2.343E+03	1.450E+03	-3.374E+02	3.481E+01	-1.344E+00
	SEGe-1	2	-1.158E+03	7.239E+02	-1.709E+02	1.789E+01	-6.999E-01
	Summed		-1.233E+02	6.828E+01	-1.522E+01	1.510E+00	-5.626E-02