New Developments in High Performance Magnetic Separation Technology for Laboratory and Industrial Applications

David Humphries, Martin Pollard

D.O.E. Joint Genome Institute/
Lawrence Berkeley National Laboratory

Emerging Magnetic Technology for Genomic, Proteomic and Bio-Medical Applications

- Hybrid 384-well magnet plates for molecular separation
- Hybrid 96-well magnet plates
- New development areas from deep well plates to single molecule manipulation
- Technology transfer and industry collaboration

Unique Technology with Broad Applicability

- New class of magnet plates developed at JGI/LBNL for high-throughput purification of biological samples in high-density microtiter plates
- Technology can selectively separate proteins, DNA and other molecules from various contaminants based solely on a magnetic field
- Magnetic structure is a unique hybrid of permanent magnet and ferromagnetic materials that produces magnetic fields significantly stronger than those of commercially available magnet plates

Hybrid Magnet with Plate Interface and 384-Well Microtiter Plate

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

High Throughput and Benchtop Use

- Hybrid magnetic structures are currently an enabling technology for sequencing approximately 2.5 Gigabases of DNA per month at the DOE Joint Genome Institute.
- Hybrid magnet plates have been developed for general bench top use in addition to automated high throughput applications.

Computer Model of Hybrid Magnetic Structure

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

High Fields and Strong Gradients

Target Species Are Magnetized

- Magnet plates work in conjunction with 0.5 to 5.0 micron magnetic beads
- Beads are attached to DNA or other targets such as proteins by means of specialized bead coatings
- Various bead types are commercially available

Applications

- Functional genomics
- Genetic sequencing
- Proteomics
- Immunological drug screening
- Automated DNA purification
- Automated protein purification
- Any magnetic bead based purification method
- Single molecule manipulation

BiomekFX Liquid Handling Robot with Four Hybrid Magnets Installed

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Close-up of Hybrids on Biomek-FX Deck

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Advantages of Hybrid Technology

- Fields are 70 to 100% stronger than that of the best commercial magnet plates measured
- Produces higher sample yields and faster processing times
- Greater holding power means more robust process
- Compatible with most 96, 384 and 1536-well standard microtiter plates

General Attributes

- These hybrid structures are energized by permanent magnets and require no external power source
- They are compact, with a footprint slightly larger than a standard microtiter plate and a thickness of approximately 1 to 1.8 inches
- Documented fabrication techniques have been developed that allow for high quantity/quality production of these structures

Performance Comparison

- Relative field strengths of five different magnet plates were compared.
- Three of the magnet plates were developed at JGI/LBNL.
- The other two magnet plates are commercially available models.
- The field strengths were measured at two heights: (a) at the magnet surface and (b) at 1 cm above the magnet surface.
- Measurements were made using a Hall effect probe.

Fields at magnet surface are 70% greater than industry 384-well magnet plates tested

Fields at magnet surface are 70 to 100% greater than industry 384-well magnet plates tested

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Fields 1 cm above the magnet are approx. 400% stronger than 384-well commercial magnet

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Drawdown Performance Comparison

- Video still sequence provides a draw-down comparison between the hybrid and a common commercial magnet plate
- 1 micron beads are suspended in a viscous solution in 384-well microtiter plates
- Non-magnetic mock-up plate is included for visual reference

Hybrid vs Commercial Plate - Initial Conditions

Drawdown After 20 Seconds

Drawdown After 30 Seconds

Drawdown After 50 Seconds

96-Well, 384-Well & 1536-Well Industry-Standard Microtiter Plates

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

New Hybrid-96 Prototype Plates In Testing

- Second generation Hybrid structures
- Optimized for higher fields and gradients
- Initial configurations for 96 well microtiter plates
- Designed for fast drawdown and beneficial distribution of targets

Hybrid 96-well Prototype

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Computer Model of 96-Well Hybrid Fields

Hybrid H96C_Trunc_v1 Model (for PANDIRA)

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Second Generation 96-well Hybrid Magnetic Structure w/ 96-well Microplate

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Second Generation 96-well Hybrid Magnetic Structure w/o Microplate

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Hybrid-96 Core Magnetic Structure

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Hybrid-96 Exploded View of Core Magnetic Structure

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Hybrid-96 Detail View of Pole Tips

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Microtiter Plate Well to Pole Relationship

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Continuing and Future Work

- Continued development of the hybrid concept for higher performance and broader applicability
- Fabrication of new, long period device for 1 and 2 ml deep well plates
- Hybrid variants for single molecule manipulation under rapid development
- Bio-medical applications of Hybrid technology in conceptual stages

Long Period Hybrid w/ 2 ml Deep Well Plate

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Long Period Hybrid w/ Upper Interface

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Long Period Hybrid w/o Upper Interface

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Long Period Hybrid – Exploded View

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

Technology Transfer Opportunities

- Hybrid magnetic technology being patented by the University of California and Lawrence Berkeley National Laboratory
- Hybrid technology available for transfer to industry through the Tech Transfer Department of LBNL
- LBNL evolving as a center for advanced magnetic R&D for genomic, proteomic and bio-medical applications
- LBNL open to collaborative development efforts as part of its tech transfer process

Advanced Fabrication Facilities at Lawrence Berkeley National Laboratory

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

State-of-the-Art High Precision Machine Tools

DEHumphries@LBL.gov - © Lawrence Berkeley National Laboratory / DOE Joint Genome Institute, 2005

This work was performed under the auspices of the US Department of Energy's Office of Science, Biological and Environmental Research Program by the University of California, Lawrence Livermore National Laboratory under Contract No. W-7405-ENG-48, Lawrence Berkeley National Laboratory under contract No. DE-AC03-76SF00098 and Los Alamos National Laboratory under contract No. W-7405-ENG-3.

LBNL-57845