ANL/ED/CP--84984

CONF-9505251--1

MODERATOR/COLLIMATOR FOR A PROTON/DEUTERON LINAC TO PRODUCE A HIGH-INTENSITY, HIGH-QUALITY THERMAL NEUTRON BEAM FOR NEUTRON RADIOGRAPHY

by

RECEIVED

JUN 19 1995

R.C. Singleterry Jr., G.R. Imel*, and G.E. McMichael**

*Argonne National Laboratory, P.O. Box 2528, Idaho Falls, ID 83403-2528 USA STI

**Argonne National Laboratory, 9700 South Cass Ave., TD/207, Argonne, IL 60439 USA

1995 Particle Accelerator Conference and International Conference on
High-Energy Accelerators
Hyatt Regency Hotel, Dallas, Texas
May 1st-5th, 1995

The submitted manuscript has been authored by a contractor of the U.S. Government under contract No. W-31-109-ENG-38. Accordingly, the U.S. Government retains a nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or allow others to do so, for U.S. Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED 65

MASTER

MODERATOR/COLLIMATOR FOR A PROTON/DEUTERON LINAC TO PRODUCE A HIGH-INTENSITY, HIGH-QUALITY THERMAL NEUTRON BEAM FOR NEUTRON RADIOGRAPHY

R.C. Singleterry Jr., G.R. Imel*, and G.E. McMichael**

*Argonne National Laboratory, P.O. Box 2528, Idaho Falls, ID 83403-2528 USA

**Argonne National Laboratory, 9700 South Cass Ave., TD/207, Argonne, IL 60439 USA

Abstract

Reactor based high resolution neutron radiography facilities are able to deliver a well-collimated ($L/D \geq 100$) thermal flux of 10^6 n/cm²·sec to an image plane. This is well in excess of that achievable with the present accelerator based systems such as sealed tube D-T sources, Van der Graaffs, small cyclotrons, or low duty factor linacs. However, continuous wave linacs can accelerate tens of milliamperes of protons to 2.5 to 4 MeV. The MCNP code has been used to analyze target/moderator configurations that could be used with Argonne's Continuous Wave Linac (ACWL). These analyses have shown that ACWL could be modified to generate a neutron beam that has a high intensity and is of high quality.

I. INTRODUCTION

Neutron Radiography (NR) has been the domain of thermal reactors until the present day. The reactor was the easiest and most economical method for the generation of a high quality, high intensity thermal neutron source. Today, reactors are becoming expensive to license and operate and are not portable. At the same time, charged particle accelerator technology is advancing, intensities are increasing, and hardware is decreasing in size. This paper investigates the use of a charged particle accelerator, the Argonne Continuous Wave Linac (ACWL) at Argonne National Laboratory in Illinois, to generate a high quality, high intensity thermal flux suitable for NR.

II. ACCELERATOR FACILITY

The original purpose of ACWL was as an accelerator development test stand for the Strategic Defense Initiative (SDI) as describe in Reference [1]. This mission has ended, so ACWL is being modified to investigate the use of accelerators in NR and Boron Neutron Capture Therapy with the current work concentrating on NR. ACWL was designed and built to accelerate negative deuteron ions in a radio frequency quadrupole accelerator in a Drift Tube Linac (DTF). The facility is being modified for water cooling versus the more expensive cryogenic cooling needed for the space environment of the SDI mission. If it is deemed that proton acceleration is needed, the particle generation equipment exists and can be fitted to ACWL readily. This analysis investigates what particle(s), particle energies, and particle currents are needed to perform NR with the intended goals.

III. NEUTRONIC ANALYSES

Various models of the new ACWL target facility were analyzed for neutronic properties with the MCNP code and ENDF/B-V continuous energy cross sections. The MCNP code system simulates many particle trajectories using a stochastic method. With these tools, analyses were performed to determine if ACWL with a suitable target could generate the beam needed to obtain the goals. The first step was to generate a source definition that modeled the neutron generation reactions. The target area and thickness were chosen to enable the target to dissipate the heat generated by the reaction. A moderator substance and configuration were designed to generate as many thermal neutrons as possible without contaminating the beam with gammas and higher energy neutrons. The beam tube was designed to transport those neutrons to the image plane and maintain its high resolution capability.

A. Source Model for the Li^7 -proton Reaction

Since MCNP cannot analyze charged particles other than electrons, the model for the source was performed separately. The reaction of interest was the $Li^7(p,n)Be^7$ reaction. Experimentally based cross section data was given in Reference [2]. The resultant energy and direction of the outgoing neutron was calculated from the two-body equations as formulated in Reference [3]. Figure 1 shows the neutron source strength as a function of proton energy and laboratory angle. Continuous slowing down theory was used to model the protons while traveling through the target. Reference [4] was used for the stopping power of hydrogen ions in the target materials.

The neutron source strength was collapsed into a source definition recognized by MCNP. To do this, the angular variable was broken into evenly distributed angle bins from 0 to 180 degrees with the probability equal to the normalized source strength. To link the angular distribution to the neutron energy distribution for that angle, the dependant source mechanism in MCNP was used. Each angle bin was made dependant to an energy range and probability as shown in Figure 2. The energy range is the maximum and minimum energies shown in Figure 2 for each angle, while the probability distribution was set equal to the normalized source strength for that angle.

B. Target Design

Exact details of the target design are not necessary for this analysis. The materials that could scatter or absorb

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

Figure 1. The Neutron Source Strength as a Function of Laboratory System Angle and Proton Energy in Lithium

Figure 2. The Neutron Energy Distribution as a Function of Laboratory System Angle and Proton Energy in Lithium

neutrons and their position are the important details modeled. The proton target is modeled as a right circular cylinder which projects enough circular surface area to the proton beam so it absorbs approximately $1 \frac{\text{kW}}{\text{cm}^2}$ of energy. The proton target is attached to a copper plate and sealed in a stainless steel 316 evacuated can that extends to the left hand edge of the model. The can is slightly larger than the proton beam to allow clear passage of the beam to the target.

C. Moderator and Reflector Design

The neutron energies from the source are between 2.5 MeV and 0.002 MeV. The goal is to have a large number of neutrons enter the beam tube with a mean energy of 0.0253 eV (thermal neutrons). The moderator must slow the neutrons down to these energies without absorbing the neutrons, spreading them through geometry effects, or having them escape. A low atomic weight, high density moderator is used to achieve the largest energy loss per collision possible to keep the geometry losses to a minimum. A reflector material is placed outside of the moderator to economize on neutrons.

The neutron source is very forward peaked. Tally surfaces were placed in the model to cover possible positions for apertures. To accelerate the MCNP calculation, gamma rays were not simulated in the models to determine placement of the aperture; however, in the final model, gamma rays were included.

D. Beam Tube Design

The objective is to obtain a collimated beam of neutrons having a thermal flux of 10^6 neutrons per $\text{cm}^2 \cdot \text{sec}$ striking the image plane with an image quality category of I as described in Reference [5]. A first order approximation for the flux entering the beam tube is

$$\phi_0 = 16\phi_I \left(\frac{L}{D} \right)^2, \quad (1)$$

where, ϕ_I is the flux at the image plane (10^6), L is the length of the beam tube, and D is the diameter of the aperture.

In order to achieve the goal of a high quality beam, an L/D of 100 is needed. From equation (1), this gives an attenuation of 1.6×10^5 , or to achieve a flux of 10^6 , the flux at the aperture must be 1.6×10^{11} neutrons per $\text{cm}^2 \cdot \text{sec}$. To absorb any stray or reflecting neutrons from or through the beam tube, a stainless steel can is lined with a thin layer of cadmium or other large cross section material (see Reference [6]). This lining keeps neutrons from entering or exiting the beam tube except at the aperture and the image and preserves the collimation of the beam.

IV. RESULTS

The results reported here are for a 2.5 MeV and a 4.0 MeV proton beam impinging on a lithium target with light and heavy water as the moderator and graphite as the reflector. The lithium target is 5 cm in radius and 0.1 cm