323

4/1/64

DP-880

MASTER

AEC RESEARCH AND DEVELOPMENT REPORT

A GAS CHROMATOGRAPH FOR DETERMINING
O₂, N₂, NO, N₂O, SO₂, AND H₂S IN STEAM

Savannah River Laboratory

Aiken, South Carolina

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.

Printed in USA. Price \$0.50
Available from the Office of Technical Services
U. S. Department of Commerce
Washington 25, D. C.

A GAS CHROMATOGRAPH FOR DETERMINING
 O_2 , N_2 , NO , N_2O , SO_2 , AND H_2S IN STEAM

by

David L. West

Approved by

H. M. Kelley, Research Manager
Analytical Chemistry Division

March 1964

E. I. DU PONT DE NEMOURS & COMPANY
EXPLOSIVES DEPARTMENT - ATOMIC ENERGY DIVISION
TECHNICAL DIVISION - SAVANNAH RIVER LABORATORY
AIKEN, SOUTH CAROLINA

CONTRACT AT (07-2) - 1 WITH THE
UNITED STATES ATOMIC ENERGY COMMISSION

ABSTRACT

A gas chromatograph was designed, assembled from existing components, and tested for the determination of O_2 , N_2 , NO , N_2O , SO_2 , and H_2S in superheated steam. In a four-column arrangement, water was stripped from the sample by a polar partition column at $160^\circ C$, and then the remaining components were resolved by the other three columns at room temperature. A sample can be analyzed automatically in 20 minutes. The minimum detectable concentration was 0.1 vol % or less for all gases except H_2S , for which the minimum was 0.3 vol %.

CONTENTS

	<u>Page</u>
Introduction	5
Summary	5
Discussion	7
Design	7
Columns and Valves	7
Auxiliary Equipment	10
Operational Sequence	11
Laboratory Tests	13
Bibliography	15

LIST OF TABLES AND FIGURES

<u>Table</u>		<u>Page</u>
I	Column Characteristics	8
II	Operational Sequence	11
III	Gas Mixtures Analyzed	13
IV	Detection Limits	14
 <u>Figure</u>		
1	The Gas Chromatograph	7
2	Flow Diagram of Four-Column Chromatograph	8
3	Inside of 160°C Cabinet	9
4	Inside of 25°C Cabinet	9
5	Control System for Chromatographic Valves	10
6	Composite Chromatogram of Off-Gases	12

A GAS CHROMATOGRAPH FOR DETERMINING O₂, N₂, NO, N₂O, SO₂, AND H₂S IN STEAM

INTRODUCTION

Analyses of the off-gases from laboratory experiments on the calcination of fission product nitrates in sulfur⁽¹⁾ were needed to follow the reactions and to measure the effects of changes in experimental conditions. The off-gas was sampled at 150°C. Within the limits of chemical compatibility, it was expected to contain major quantities of steam, SO₂, NO, and NO₂, and lesser quantities of SO₃, N₂O, N₂, and possibly H₂S and O₂.

Literature references on work with nitrogen compounds and sulfur compounds indicated that gas chromatography is readily adaptable to the determination of each of these gases except SO₃ and NO₂. O₂, N₂, and NO are resolvable on a molecular sieve column,^(2,3) and N₂O is separable from these gases on a charcoal column.⁽³⁾ SO₂, H₂O, and H₂S are separable by a dual-column arrangement consisting of a column of 10% polyethylene glycol on "Teflon"* and a column of 30% polyethylene glycol on firebrick.⁽⁴⁾ An indirect gas chromatographic method for determining NO₂⁽⁵⁾ is more complex and less reliable than a direct photometric method that was chosen. SO₃ is too corrosive to be determined chromatographically.

SUMMARY

A semiautomatic four-column gas chromatograph was designed for the determination of O₂, N₂, NO, N₂O, SO₂, and H₂S in the off-gas from the calcination of radioactive waste in sulfur. In a four-column arrangement, water was stripped from the sample by a polar partition column at 160°C, and then the remaining components were resolved by the other three columns at room temperature.

The chromatograph operated satisfactorily on 300 samples of wet and dry mixtures of pure gases. A sample could be analyzed automatically in 20 minutes. No column deterioration or corrosion was detected. The minimum detectable concentration was 0.1 vol % or less for all gases except H₂S, for which the minimum was 0.3 vol %.

*Registered trademark of E. I. du Pont de Nemours & Co. for tetrafluoroethylene resins.

**THIS PAGE
WAS INTENTIONALLY
LEFT BLANK**

DISCUSSION

DESIGN

The gas chromatograph shown in Figure 1 was designed and built for the determination of O_2 , N_2 , NO , N_2O , SO_2 , and H_2S in the off-gas. The chromatograph is a multicolumn unit of modular design and uses helium as the carrier gas.

FIG. 1 THE GAS CHROMATOGRAPH

Columns and Valves

A four-column arrangement was selected to remove water and to resolve the desired components. The determination of water, though possible by gas chromatography, was omitted to reduce the complexity of the chromatograph. A flow diagram of this arrangement is shown in Figure 2, and the columns are described in Table I. The sample valve, Column 1, and the backflush valve are in the $160^{\circ}C$ cabinet (see Figure 3). This thermostatted oven is a radically modified version of the one supplied with a Consolidated Laboratory Chromatograph Type 26-201.

Both dual-column valves, the detector, the valve control system, and Columns 2, 3, and 4 are in the $25^{\circ}C$ cabinet (see Figure 4). This cabinet is not thermostatted, but is at room temperature, which is maintained at $25^{\circ}C$. The detector, a thermal conductivity type, is from the Consolidated unit.

FIG. 2 FLOW DIAGRAM OF FOUR-COLUMN CHROMATOGRAPH

TABLE I

Column Characteristics

Column Number	Packing Material	Length, inches ^(a)	Temperature, °C
1	5% "Carbowax" ^(b) 1500 on 110-120 mesh "Teflon" 6	24	160
2	5% "Carbowax" ^(b) 1500 on 110-120 mesh "Teflon" 6	24	25
3	4% 2,5-hexanedione on charcoal	3	25
4	5A molecular sieve, 30-60 mesh	48	25

(a) The tubing for all columns was type 316 stainless steel with 0.25-inch OD and 0.035-inch wall thickness.

(b) Registered trademark of Union Carbide Corporation for polyethylene glycols.

DPSTF-1-7042

FIG. 3 INSIDE OF 160°C CABINET

DPSTF-1-7044

FIG. 4 INSIDE OF 25°C CABINET

In the valve control system (see Figure 5) a motor-driven cam assembly operates the solenoid valves that control the air flow to the chromatographic valves. The pneumatically operated chromatographic valves are Greenbrier Type C-4 valves connected in the configurations shown in Figure 2. The sample and backflush valves are made of type 316 stainless steel and have "Rulon"* O-rings. The dual-column valves are made of type 303 stainless steel and have "Viton"** A O-rings. The sample loop is a piece of 1.4-mm-ID stainless steel tubing attached to the sampling valve. The loop has a volume of approximately 1 cc.

All the tube fittings are "Swagelok"*** brand, and both the fittings and the tubing are made of type 316 stainless steel. The threaded valve fittings are sealed with "Teflon" tape.

FIG. 5 CONTROL SYSTEM FOR CHROMATOGRAPHIC VALVES

Auxiliary Equipment

The control unit from the Consolidated Chromatograph is used to control the detector and to control and measure the flow of carrier gas. Electric power for this unit is supplied by a Sorenson Model E-6/2-5 Nobatron. The output signal from the detectors is recorded by a Brown Model 153 X recorder with a 1-mv range and a 1-second pen speed.

*Registered trademark of Dixon Corporation for glass-filled "Teflon."

**Registered trademark of E. I. du Pont de Nemours & Co. for a fluoroelastomer.

***Registered trademark of Crawford Fitting Co.

OPERATIONAL SEQUENCE

The chromatograph analyzes the sample stream repetitively on a 20-minute cycle. The operational sequence, summarized in Table II, is arranged so that the components are separated on successive columns. Each separated component is shunted to the detector before it reaches a column with which the component is incompatible. Column 1 dries the sample, Column 2 separates SO_2 from the sample, Column 3 separates H_2S and N_2O from the sample and from each other, and, finally, Column 4 separates O_2 , N_2 , and NO from each other.

TABLE II

Operational Sequence

Time (a)		Valve	Operation
Minutes	Seconds		
0	0	Sample	Bypass sample stream. Connect sample loop to chromatograph.
0	17	Sample	Return sample loop to sample stream.
0	30	Backflush	Flush water to vent. Dried sample is in Column 2.
1	18	Dual-Column 1 and 2	Direct SO_2 to detector. H_2S and N_2O are in Column 3. O_2 , N_2 , and NO are in Column 4.
4	50	Dual-Column 1	Restore flow to Column 3. H_2S and N_2O flow to detector.
14	38	Dual-Column 2	Restore flow to Column 4. O_2 , N_2 , and NO flow to detector.
19	35	Backflush	End backflush. Set chromatograph for next sample.

(a) At time 0:00 all valves are in Position 1 of Figure 2.

The composite chromatogram in Figure 6 depicts the peak shape and elution time for each component. This figure was assembled from chromatograms of the individual components, rather than measured directly, because certain combinations of the components are incompatible, e.g., NO and O₂, H₂S and SO₂.

FIG. 6 COMPOSITE CHROMATOGRAM OF OFF-GASES

LABORATORY TESTS

Laboratory tests with pure gases and the synthetic mixtures listed in Table III demonstrated the operability of this chromatograph on streams that contain O₂, N₂, NO, N₂O, SO₂, H₂S, or H₂O at 150°C. Approximately 300 samples of mixtures (including 200 dry and 100 wet) were analyzed. Peaks with the shapes shown in Figure 6 were obtained with pure gases, dry mixtures, and wet mixtures. The tests were designed to detect both column deterioration and corrosion of the metallic components. The tests were completely successful, and there was no sign of either type of failure. The effects of SO₃ or NO₂ on this chromatograph were not determined. If these components cause either type of failure, they can probably be removed from the stream before sampling.

TABLE III

Gas Mixtures Analyzed

<u>Mixture</u>	<u>Type</u>	<u>Approximate Composition, vol %</u>	<u>Approximate No. of Analyses</u>
1	Dry	50% N ₂ O, 10% O ₂ , 40% N ₂	50
2	Dry	50% H ₂ S, 10% O ₂ , 40% N ₂	50
3	Dry	50% SO ₂ , 10% O ₂ , 40% N ₂	50
4	Dry	20% O ₂ , 80% N ₂	50
5	Dry	10% NO, 90% He	10 ^(a)
6	Wet	Mixture 1 plus steam ^(b)	25
7	Wet	Mixture 2 plus steam	25
8	Wet	Mixture 3 plus steam	25
9	Wet	Mixture 4 plus steam	25

(a) Analysis limited by shortage of NO

(b) Gas bubbled through 100°C water and then piped to analyzer. Pipe maintained at 150°C. No analysis made with wet Mixture 5 due to shortage of NO.

The minimum detectable concentration for each of the components determined is shown in Table IV. These are concentrations that would give a peak height of 10% of full scale on the most sensitive range of the detector, i.e., an 0.1-mv signal. This chromatograph was not designed for high sensitivity. Substitution of a low-noise detector and improvements in the thermostat and flow controls would increase the sensitivity at least tenfold. The precision of the method was not determined, but on the basis of previous experience, coefficients of variation are expected to be less than 10%.

TABLE IV

Detection Limits

<u>Component</u>	<u>Minimum Detectable Concentration,</u>	
	<u>vol %</u>	
SO ₂	0.06	
O ₂	0.08	
N ₂ O	0.1	
NO	0.1	
N ₂	0.1	
H ₂ S	0.3	

BIBLIOGRAPHY

1. Davis, M. W., Jr., E. S. Occhipinti, and C. B. Goodlett. Calcination of Radioactive Waste in Molten Sulfur. E. I. du Pont de Nemours & Co., Savannah River Laboratory, Aiken, S. C. USAEC Report DP-713 (1962).
2. Graven, W. M. "Kinetics of the Decomposition of N_2O at High Temperature." J. Am. Chem. Soc. 81, 6190-92 (1959).
3. Smith, D. H. and F. E. Clark. "Some Useful Techniques and Accessories for Adaptation of the Gas Chromatograph to Soil Nitrogen Studies." Soil Sci. Soc. Am. Proc. 24, 111-15 (1960).
4. Villalobos, R. "Applications of 'Teflon' as a Solid Support in Industrial Gas Chromatography." Summer Instrument-Automation Conference and Exhibit. Instrument Society of America Preprint 15-SF60 (1960).
5. Greene, S. A. and H. Pust. "Determination of Nitrogen Dioxide by Gas-Solid Chromatography." Anal. Chem. 30, 1039-40 (1958).