Presented at the Radiological Society
of North America, Inc., November 29-
December 4, 1970, Chicago, Ill.

UCRL-20190
Preprint

CONF-701103--7

MASTER

MEASUREMENT OF OXYGEN EFFECT AND
BIOLOGICAL EFFECTIVENESS OF A 910-MeV
HELIUM ION BEAM USING CULTURED CELLS (T41)

Mudundi R. Raju, Madhvanath Gnanapurani,
Bambino Martins, Jerry Howard, and John T. Lyman

November 1970

AEC Contract No. W-7405-eng-48

UCRL

LAWRENCE RADIATION LABORATORY
UNIVERSITY of CALIFORNIA BERKELEY

UCRL-20190

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency Thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

DISCLAIMER

Portions of this document may be illegible in electronic image products. Images are produced from the best available original document.

PAGES i to ii
WERE INTENTIONALLY
LEFT BLANK

MEASUREMENT OF OXYGEN EFFECT AND BIOLOGICAL
EFFECTIVENESS OF A 910-MeV HELIUM ION BEAM
USING CULTURED CELLS (T-1)

MUDUNDI R. RAJU, D.Sc., MADHVANATH GNANAPURANI, M.S.
BAMBINO MARTINS, M. Sc., JERRY HOWARD, M.S.,
and JOHN T. LYMAN, Ph.D.

ABSTRACT

The depth-dose distribution of a 910-MeV monoenergetic helium ion beam is modified by using a ridge filter in the beam path so that the region of maximum dose is uniform over a distance of about 6 cm of water. The biological effectiveness and oxygen enhancement ratio (OER) at different depths within a rectangular Lucite phantom are measured by using human kidney cells (T-1). The results indicate that the biological effectiveness at the broad peak region is about 1.3 to 1.4 compared with that at the beam entrance, and the OER is found to be about 1.7 to 1.9. This significant reduction in OER even when the beam is modified to cover 6 cm of tissue may be of radiotherapeutic interest in treatment of deep seated tumors containing hypoxic cells.

NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

Principal Author:

Mudundi R. Raju, D.Sc.

Bldg. 29, Room 229A

Lawrence Radiation Laboratory

Berkeley, CA 94720

Key Words:

Radiotherapy

Relative biological effectiveness

Oxygen enhancement ratio

Helium ion beam

Tissue culture

Ridge filter

INTRODUCTION

In radiation therapy, ideally one would like to deliver a tumorcidal dose to the tumor; however, in practice one is limited by normal-tissue injury. Radiations and techniques that help in delivering more dose to the tumor with minimum dose to the surrounding normal tissue could improve the cure rates and reduce the complications due to normal-tissue injury. Significant progress has been made by employing penetrating radiations. However, the use of heavy charged particles such as protons and helium ions is not very extensive, although such radiations are promising in improving radiotherapy results in principle and have been available at a few places for many years. This is probably because these facilities (cyclotrons) are not in radiotherapy departments.

The therapeutic use of the 910-MeV helium ion beam from the 184-in. -synchrocyclotron at Berkeley has been mainly for pituitary irradiations of patients with acromegaly, diabetic retinopathy, Cushing's disease, or breast cancer (1). When the narrow Bragg peak is broadened by means of a variable absorber in the beam, it is possible to use this beam for other radiotherapeutic applications. Since, in such cases, the dose contribution due to high-LET components at the broadened Bragg-peak region is small, it was felt that such a beam may not be useful in overcoming the oxygen effect considerably (2). However, the experimental determination of the oxygen enhancement ratio of 1.6 for

14-MeV neutrons indicates that a small fraction of dose due to high-LET alpha particles produced by 14-MeV neutrons in tissue is mainly responsible for reducing the OER (3, 4).

It is of interest to find out how a small fraction of dose due to high-LET components (low-energy alpha particles) at the broadened Bragg-peak region would affect the radiobiological properties of therapeutic interest, such as biological effectiveness and oxygen enhancement ratio. Such measurements will be helpful in assessing the therapeutic capabilities of this beam.

Our preliminary measurements of oxygen effect at the broadened Bragg peak region over a depth of 3 cm indicated significant reduction in oxygen enhancement ratio (5). In this investigation the 910-MeV helium ion beam is modified so that the dose at the peak region is uniform over 6 cm of water. The biological effectiveness and OER over this broad-peak region and at the beam entrance are measured.

MATERIALS AND METHODS

Physical

The width of the Bragg peak of a monoenergetic charged particle beam is small when compared with typical tumor sizes. One technique for modifying the depth-dose distribution of monoenergetic charged particles to uniformly irradiate the treatment volume is to use a ridge filter (6, 7).

The ridge filter consists of a series of similar units, usually made of a dense metal such as copper. These units are placed side by side to form a composite filter whose cross section is larger than the beam area. Each unit is essentially a stepwise variable-thickness absorber. The width of an individual step determines the relative intensity, and the total thickness of the step determines the residual energy. The use of a ridge filter permits the simultaneous superposition of the depth-dose distributions of all the components required to produce the desired depth-dose distribution. As more components are used to approximate the desired distribution, the number of steps increases and the cross section of the individual units more closely approximates a smooth curve.

The 910 MeV helium ion beam has a range of about 32 cm in water; this energy is much higher than is necessary for radiotherapeutic applications. Such high-energy heavy charged-particle beams give a lower ratio of dose at the peak to the entrance because of multiple scattering effects and the loss of the particles before they stop. The beam is first degraded by a 3.2-cm copper absorber. The residual energy of the beam after passing through copper is 500 MeV. This residual beam is further modified by using a ridge filter so that the peak of the depth-dose distribution is broad enough to cover 6 cm of water. In a Lucite phantom this is equivalent to about 5 cm ($6\text{g}/\text{cm}^2$ of Lucite).

The design of the ridge filter is optimized by the use of a computer program which can translate either an experimental or

a theoretical narrow Bragg curve to the required curve with a broadened peak region. The output of the computer program includes the necessary information for punching a paper tape, which can be used with a paper-tape-controlled milling machine to make the required filter. The experimental measurements of the modified depth-dose distribution agree very well with the computer-predicted dose distribution (8). The computer printout of the ridge filter used in this investigation is shown in Fig. 1, and the experimental determination of the depth-dose distribution of helium ion beam is shown in Fig. 2.

Biological

Human kidney cells (T-1) were used in this investigation. The tissue culture techniques employed have already been described(9,10). Instead of feeder cells, enriched medium containing 15% fetal calf serum was used to improve the plating efficiency. Cells in early logarithmic growth phase were plated in 35-mm plastic dishes in appropriate numbers so that surviving colonies would be about 100 in each dish. The dishes were placed in the incubator at 37°C for 4 to 7 hours before exposure to radiation.

A sample loading wheel was used for mounting the dishes for irradiation(11). The medium was removed from the dishes before they were mounted in the wheel. Compressed air or nitrogen with 5% carbon dioxide saturated with water vapor was admitted into the wheel and circulated over the dishes during exposure. In the case of exposures in nitrogen, the cells were

pregassed for at least 5 minutes before exposure to radiation. Dishes containing cells were exposed to different doses of radiation at three positions marked 1, 2, and 3 in Fig. 2. Three to four dishes were exposed for each dose point. Twelve to fifteen days after exposure the colonies were stained with methylene blue. All the visible colonies were counted and the percent survival calculated.

RESULTS AND DISCUSSION

Figure 3 shows the survival curves obtained for exposures of the three positions in a Lucite phantom. Biological effectiveness compared with that of entrance position (marked 1 in Fig. 2) and OER were calculated at 10% survival.

The biological effect at entrance is very similar to that of conventional radiation. The biological effectiveness at positions 2 and 3 compared with position 1 is 1.3 and 1.4 respectively. The OER values at positions 1, 2, and 3 are 2.5, 1.8, and 1.7 respectively. These preliminary results indicate that there is a significant reduction in OER in the broad peak region. This significant reduction in OER even when the peak is broadened to cover 6 cm of water is rather unexpected, in the same way an OER of 1.6 was for 14-MeV neutrons.

The trend of variation of RBE and OER over the peak region can be expected from variation of LET, which increases with depth. Early experiments with anoxic mammalian tumor cells studied in vivo indicate that dose response was significantly

altered at the broadened Bragg peak of a 910-MeV helium ion beam, when compared with that at the entrance (12).

A high-energy helium ion beam has a radiotherapeutic application similar to that of high-energy protons because of its excellent depth-dose distribution characteristics in the treatment of deep-seated tumors located near sensitive and vital structures. Further, the significant reduction in OER indicates an additional advantage of high-energy helium ion beams in treating tumors containing hypoxic cells. The anoxic gain factor (damage to anoxic cells when compared with oxygenated cells) for helium ion beam is greater than conventional radiation, but slightly less than that for fast neutrons. Unlike with neutrons, this gain factor is in the region of interest only and the type of damage to surrounding normal tissue is similar to that of conventional radiation. The anoxic gain factor for neutrons when compared with conventional radiation gets reduced appreciably for small fractions of doses (13). This may possibly be true for helium ion beams. However, with the excellent depth-dose distribution and beam shaping that can be obtained for helium ion beam, it is possible to deliver large single doses to the tumor region and spare the normal surrounding tissue. Thus, helium ion beam may be very effective in radiotherapy where large doses of conventional radiation have been found to be useful.

ACKNOWLEDGMENTS

One author (M. R. Raju) is grateful to Dr. G. W. Barendsen for introducing him to tissue culture work. We are grateful to Dr. J. H. Lawrence, Dr. C. A. Tobias and Dr. C. Richman for their encouragement.

M. R. Raju and M. Gnanapurani are also associated with University of Texas at Dallas.

This work was supported by the U. S. Atomic Energy Commission.

REFERENCES

1. Lawrence JH, Tobias CA: Heavy particles in medicine. Progr At Med 1:127-146, 1965
2. Fowler JF: Neutrons in radiotherapy: Slow neutrons, fast neutrons and other heavy particles. [In] Biological Effects of Neutron and Proton Irradiation. Vienna, IAEA, 1964, Vol 4, pp 185-214
3. Neary GJ, Savage JRK: Oxygen effect with 14 MeV neutrons. Nature 204:197, 1964.
4. Barendsen GW, Broerse JJ: Dependence of the oxygen effect on the energy of fast neutrons. Nature 212:722-724, 1966
5. Raju MR, Gnanapurani M, Madhvanath U, Howard J, Lyman JT: Relative biologic effectiveness and oxygen enhancement ratio at various depths of a 910 MeV helium ion beam. Acta Radiologica (in press)

6. Karlsson BG: Methoden zur Berechnung und Erziehung Einiger für die Tiefentherapie mit Hochenergetischen Protonen günstiger Dosisverteilungen. Strahlentherapie 124:481, 1964.

7. Raju MR, Lyman JT, Brustad T, Tobias CA: Heavy charged-particle beams. [In] Radiation Dosimetry ed by F H Attix and E Tochilin. New York and London, Academic Press, 1969, Vol III, pp 151-199

8. Lyman JT, Lawrence Radiation Laboratory, Berkeley, unpublished

9. Barendsen GW, Beusker TLJ, Vergroesen AJ, Budke L: Effects of different ionizing radiations on human cells in tissue culture. II. Biological experiments. Radiation Res 13:841-849, 1960

10. Todd PW: Reversible and irreversible effects of ionizing radiations on the reproductive integrity of mammalian cells cultured in vitro (Ph. D. thesis), Lawrence Radiation Laboratory Report UCRL-11614 (1964)

11. Todd PW, Lyman JT, Armer RA, Skarsgard LD, Deering RA: Dosimetry and apparatus for heavy ion irradiation of mammalian cells in vitro. Radiation Res. 34:1-23, 1968

12. Berry RJ, Andrews JR: The response of mammalian tumor cells in vivo to radiations of differing ionizing densities (LET). Ann. N Y Acad Sci 114:48-59, 1964

13. Fowler JF: Fast neutron therapy--Physical and biological consideration. [In] Modern Trends in Radiotherapy, ed by Thomas J Deeley and Constance A P Wood. London, Butterworths, 1967, pp 145-170

FIGURE LEGENDS

Fig. 1. Ridge filter.

Fig. 2. Modified depth-dose distribution of a 910-MeV helium ion beam, showing the three exposure positions.

Fig. 3. Survival curves for T-1 cells exposed under aerobic and anoxic conditions at the three positions marked 1, 2, and 3 in Fig. 2.

XBL 7011-7024

Fig. 1.

DBL 7011 5964

Fig. 2

DBL 7011 5963

Fig. 3

LEGAL NOTICE

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or*
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.*

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.