High resolution studies of the origins of polyatomic ions in inductively coupled plasma – mass spectrometry

by

Jill Wisnewski Ferguson

A dissertation submitted to the graduate faculty
in partial fulfillment of the requirements for the degree of

DOCTOR OF PHILOSOPHY

Major: Analytical Chemistry

Program of Study Committee:

R. S. Houk, Major Professor

Mark S. Gordon

David Laird

Shang-Yi Victor Lin

Marc D. Porter

Iowa State University

Ames, Iowa

2006

Graduate College
Iowa State University

This is to certify that the doctoral dissertation of

Jill Wisnewski Ferguson

has met the dissertation requirements of Iowa State University

A handwritten signature in dark ink, appearing to read "D. A. Hank", is written over a horizontal line.

Major Professor

For the Major Program

TABLE OF CONTENTS

CHAPTER 1. INTRODUCTION	1
General Introduction	1
Method Review	2
Summary	7
Dissertation Organization	8
References	8
CHAPTER 2. HIGH RESOLUTION STUDIES OF THE ORIGINS OF POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS SPECTROMETRY	
PART I. IDENTIFICATION METHODS AND EFFECTS OF NEUTRAL GAS DENSITY ASSUMPTIONS, EXTRACTION VOLTAGE, AND CONE MATERIAL	10
Abstract	10
Introduction	10
Theory	12
Experimental	14
Results and Discussion	16
Conclusions	27
Acknowledgements	28
References	28
Tables	34
Figures	38
CHAPTER 3. HIGH RESOLUTION STUDIES OF THE ORIGINS OF POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS SPECTROMETRY	
PART II. NOVEL CALCULATIONS OF UNUSUAL IONS	48
Abstract	48
Introduction	49
Theory	49
Experimental	51
Results	52
Discussion	57
Conclusions	60
Acknowledgements	60
References	61
Tables	63
Figures	69
CHAPTER 4. HIGH RESOLUTION STUDIES OF THE ORIGINS OF POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS SPECTROMETRY	
PART III. CrO_xH_y^+ IONS	70
Abstract	70
Introduction	70

Experimental	71
Results and Discussion	73
Conclusions	78
Acknowledgements	78
References	79
Tables	82
Figures	85
 CHAPTER 5. GENERAL CONCLUSIONS	 91
 ACKNOWLEDGEMENTS	 93

CHAPTER 1. INTRODUCTION

General Introduction

The inductively coupled plasma (ICP) is an atmospheric pressure ionization source. Traditionally, the plasma is sampled via a sampler cone. A supersonic jet develops behind the sampler, and this region is pumped down to a pressure of approximately one Torr. A skimmer cone is located inside this zone of silence to transmit ions into the mass spectrometer. The position of the sampler and skimmer cones relative to the initial radiation and normal analytical zones of the plasma is key to optimizing the useful analytical signal [1].

The ICP both atomizes and ionizes the sample. Polyatomic ions form through ion-molecule interactions either in the ICP or during ion extraction [1]. Common polyatomic ions that inhibit analysis include metal oxides (MO^+), adducts with argon, the gas most commonly used to make up the plasma, and hydride species. While high resolution devices can separate many analytes from common interferences, this is done at great cost in ion transmission efficiency – a loss of 99% when using high versus low resolution on the same instrument [2]. Simple quadrupole devices, which make up the bulk of ICP-MS instruments in existence, do not present this option. Therefore, if the source of polyatomic interferences can be determined and then manipulated, this could potentially improve the figures of merit on all ICP-MS devices, not just the high resolution devices often utilized to study polyatomic interferences.

Method Review

Collision Cells

Many techniques have been considered to control the development of polyatomic interferences. Chemical reactions can be used to the scientist's advantage through the use of collision or dynamic reaction cells. Hydrogen, methane, nitrogen, and xenon are gases that are commonly used to reduce the appearance of argon adducts and metal oxides. For example, using nitrogen as the collision gas reduces the $^{40}\text{Ar}^{16}\text{O}^+$ signal, but increases the abundance of $^{40}\text{Ar}^{14}\text{N}^+$ and other nitrogen based interferences [2, 3]. This becomes very limiting as users employ the multi-element capabilities for which ICP-MS is well known. Therefore, the use of collision cells is not a complete treatment for the polyatomic problem. Collision cells are a useful tool, but only under certain circumstances.

Cool Plasma

Another alternative is using a "cool" plasma, i.e. a lower RF power, higher nebulizer gas flow rate, and sampling position farther upstream. This technique is particularly helpful when Ar^+ , ArH^+ , or ArO^+ are an interference, as in the analyses of potassium, calcium, and iron. However, cool plasmas inefficiently ionize elements with a high ionization potential, such as selenium. This technique also tends to increase the abundance of refractory oxides and analyte hydrides [2, 3]. Cool plasmas are suitable as a sample specific tool for controlling polyatomic interferences.

Solvent Removal

Reducing the solvent load introduced to the plasma can reduce the abundance of many polyatomic ions. Minnich and Houk compared conventional, cryogenic, and

membrane desolvation, including combinations of the techniques [4]. MO^+ , MOH^+ , and MH^+ signals were monitored, as well as polyatomic ions containing chlorine. It was found that cryogenic desolvation removes more MO^+ than membrane desolvation. However, this effect is not additive when the two techniques are combined. The combination of cryogenic and membrane desolvation does reduce MH^+ to near background levels, but this comes at a price of reduced sensitivity, roughly to one-half that of the other methods employed singularly. Cryogenic desolvation also reduced the prevalence of chlorine polyatomic ions [4]. It is important to note that this technique is less convenient than other methods. In general, desolvation techniques often suffer from greater memory effects as well.

Alternate Stability Regions

Alternate stability regions provide improved mass resolution for quadrupole devices, which in turn allows for separation of some interferences from analytes of interest. However, the resolutions reached are still inferior to those provided by a sector field device. In addition, operation in alternate stability regions results in reduced sensitivity and increased background levels [2]. The higher RF power required for the use of alternate stability regions also limits the feasibility of the technique. The use of alternate stability regions is not a panacea for the polyatomic ion issue.

Cone Composition and Geometry

Another way that researchers have tried to reduce polyatomic ion formation is through modification of the sampler and/or skimmer cones. Helal and coworkers have studied the influence of skimmer materials on the formation of polyatomic ions [5]. While using a copper skimmer cone plated with gold, they monitored polyatomics that would be

expected to form, such as ArX^+ , XO^+ , and XN^+ , where X indicates Cu or Au. The signal from each polyatomic ion was compared for Cu and Au. It was found that the gold was more likely to create the expected interferences than copper, and as the gold was stripped off the outer surface of the cone, the gold interferences were reduced. Over the course of forty minutes, the ratios of the copper and gold interferences neared one [5]. This result shows the impact of cone material in polyatomic ion formation.

Other variables that can impact the formation of polyatomic interferences are aperture size, cone spacing, and angle design. The aperture size should be maximized to allow for the greatest analytical signal and to avoid clogging, however the size is limited by the pumping capability, as the interface region should maintain a pressure of about 1-4 Torr [6].

Typically, this maximum size is reached in the 0.7 to 1.0 mm range. Oxide formation increases with a decrease in sampler orifice diameter, indicating a boundary layer around the orifice [7]. The spacing between the sampler and skimmer also impacts oxide formation. With an increased distance between the two cones, oxide signal decreases relative to analyte signal. However, the reduction in analyte sensitivity can be problematic [8].

A group headed by Jarvis studied three popular nickel skimmer cone angle designs using two different ICP-MS instruments [3]. In all cases, the instrument optics and gas flow rates were optimized to result in the maximum atomic ion signal. Common polyatomics that were monitored included $^{14}\text{N}_2^+$, $^{12}\text{C}^{16}\text{O}_2^+$, and $^{40}\text{Ar}^{38}\text{Ar}^+$. The “PolySkim” design, with an internal angle smaller than 45° , external angle greater than 60° , and an orifice diameter of 0.7 mm demonstrated the greatest reduction in $^{40}\text{Ar}^{16}\text{O}^+$ and the other polyatomics monitored. This design had a smaller internal angle and intermediate external angle when compared with the other two configurations tested. The “Polyskim” model provided a more stable signal,

with better signal to noise characteristics than the other designs. However, there was a decrease in absolute sensitivity noted [3].

Sampling Position and Concomitant Elements

The position at which the ICP is sampled heavily influences the makeup of the mass spectrum. Shifts in ionization equilibria can be seen as one scans radially and axially across the plasma. It has been demonstrated that oxide fractions $[MO^+ / (M^+ + MO^+)]$ can be used as an indicator of these shifts in ionization equilibrium because this ratio is not changed by variations in lens potentials, space-charge effects, or plasma conditions [9]. Beauchemin has studied this and the effect of concomitant elements in relation to the polyatomic interferences problem [10].

Beauchemin observed analytes (La^+ and Ce^+) and their oxides while moving the plasma either across or away from the sampling interface, both with and without the presence of concomitant elements. LaO^+ and CeO^+ are strongly bound, and can cause substantial interferences in the analysis of heavier rare earth elements. She found that the oxides were present at a sampling position farther upstream than their atomic ions. This is because the plasma temperature in this region is cooler [10]. Cooler plasma conditions increase the metal oxides present [2]. As one samples at greater distances, the oxides are dissociated, and the atomic ion signal initially increases. The La^+ or Ce^+ signal then declined as all the oxides were broken up. The same behavior was noted when the test was repeated in a 0.02 molar sodium matrix [10].

While this same oxide dissociation activity was seen in the presence of concomitant elements, the sampling positions at which this dissociation took place were shifted to positions closer to the load coil (smaller sampling distances). This shift was ascribed to a

shift in the ionization equilibrium as discussed previously. This shift in equilibrium is most evident at shorter sampling distances where the electron prevalence is changed. This shows that while similar trends were noted in the presence of additional elements, the actual distances for best atomic sampling differed [10]. This again leads us to the specificity required for each sample analysis in ICP-MS. This work shows the importance of treating polyatomic interferences in a case by case manner. The more one knows about a sample, the faster the optimal experimental parameters for a specific analysis can be reached. Polyatomic interferences perform differently depending on additional species in solution, instrument, and experimental parameters.

Mathematical Corrections

Computational methods of correcting for interferences are beneficial for the lab technician, often requiring no adjustment of current analytical protocol or of the instrument itself. In simple cases on a quadrupole device, one could correct for the presence of $^{40}\text{Ar}^{12}\text{C}^+$ interference on $^{52}\text{Cr}^+$ by monitoring the signal at m/z 13 and 52. m/z 13 is monitored to detect carbon rather than m/z 12 because the $^{12}\text{C}^+$ signal would overload the detector, especially if the carbon is mainly coming from carbonate in solution. Through repeated trials, the signal ratio $^{40}\text{Ar}^{12}\text{C}^+ / ^{13}\text{C}^+$ can be estimated. By repeatedly analyzing a blank that does not contain chromium, but has similar carbon content to the sample, one can establish a simple correction factor for m/z 52, where $^{40}\text{Ar}^{12}\text{C}^+$ would ordinarily prevent analysis [11]. This simple computational method requires repeated tests and knowledge of the sample and its matrix. The correction factor is specific for each analytical run and would need to be repeated from day to day for superior accuracy. These corrections break down when the signal from the interferent is close to or greater than the analyte signal. Of course, it is

always preferable to remove the source of error, the interference, rather than correct for its estimated abundance.

Much more advanced computational methods exist, often employing principal components analysis. Software has been developed that automates much of this process. Existing multicomponents analysis (MCA) software allows a model to be initialized from a full mass scan ($m/z = 6-238$.) This program takes into account shot noise and plasma flicker characteristics. It then excludes interferences from the model when their signal is below the detection limits projected for the sample matrix. The model is then used to give a picture of what the “true” spectrum should look like [11]. Computational methods for removing polyatomic interferences are only as good as the models they develop. Eliminating the interferences themselves is generally preferred over regulating their presence with statistical models.

Summary

Researchers currently have numerous ways to address the interferences caused by polyatomic ions. All of these methods are limited by their poorer sensitivity, creation of other interferences, or the uncertainty of mathematical corrections. While there are volumes of ICP-MS research dedicated to methods of reducing polyatomic interferences, few have studied how they may develop in the first place. Since the ICP is thought of as a complete atomization and ionization source at a temperature of approximately 7000 K, the presence of these unusual ions is remarkable. Fundamental research aimed at understanding the formation or dissociation of these ions, as well as determining the location of these processes, may be the basis for truly being able to successfully influence the prevalence of polyatomic ions in spectra.

Dissertation Organization

This dissertation is organized into chapters. Chapters 2, 3, and 4 are each complete scientific journal articles prepared for submission to *Spectrochimica Acta Part B*. Chapter 2 is part one in a three part series concerning the study of polyatomic ions in ICP-MS using a high resolution device. This chapter discusses ion identification techniques as well as the effect of instrumental parameters on the appearance of common polyatomic ions such as H_2O^+ , ArO^+ , and NO^+ . Chapter 3 is the second paper in the series. It discusses polyatomic ions not often studied, such as N_2H^+ and H_xCO^+ , where $x=1-3$. Novel calculations of the energy states of these ions have been done by members of the Gordon group, and this collaboration brings to light some interesting questions about polyatomic ions. Chapter 4 completes the current series, and provides insight into unusual spectral interferences seen in the analysis of palladium. These curious polyatomic ions contain chromium and challenge the traditional position held by most about the relative abundance of polyatomic ions. General conclusions and the directions of future research are summarized in chapter 5.

References

- [1] H. Niu and R. S. Houk, Fundamental aspects of ion extraction in inductively coupled plasma mass spectrometry, *Spectrochim. Acta Part B* 51 (1996) 779-815.
- [2] F. Vanhaecke and L. Moens, Recent trends in trace element determination and speciation using inductively coupled plasma mass spectrometry, *Fresenius J. Anal. Chem.* 364 (1999) 440-451.
- [3] K. E. Jarvis, P. Mason, T. Platzner, J. G. Williams, Critical assessment of the effects of skimmer cone geometry on spectroscopic and non-spectroscopic interferences in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 13 (1998) 689-696.

- [4] M. G. Minnich, R. S. Houk, Comparison of cryogenic and membrane desolvation for attenuation of oxide, hydride and hydroxide ions and ions containing chlorine in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 13 (1998) 167-174.
- [5] N. F. Zahran, A. I. Helal, M. A. Amr, A. Abdel-Hafiez, H. T. Mohsen, Formation of polyatomic ions from the skimmer cone in the inductively coupled plasma mass spectrometry, *Int. J. Mass Spectrom.* 226 (2003) 271-278.
- [6] A. Montaser, ed. *Inductively Coupled Plasma Mass Spectrometry*, New York: Wiley, 1998, 425.
- [7] M. A. Vaughan, G. Horlick, Effect of sampler and skimmer orifice size on analyte and analyte oxide signals in inductively coupled plasma – mass spectrometry, *Spectrochim. Acta Part B* 45 (1990) 1289-1299.
- [8] J. W. H. Lam, G. Horlick, Effects of sampler – skimmer separation in inductively coupled plasma – mass spectrometry, *Spectrochim. Acta Part B* 45 (1990) 1327-1338.
- [9] S. D. Tanner, ICP-MS tectonics: moving the atomic and oxide ion ‘mountains’, paper F1. *Proceedings of the 1994 Winter Conference on Plasma Spectrochemistry*, San Diego, January 1994.
- [10] M. M. Fraser and D. Beauchemin, Effect of concomitant elements on the distribution of ions in inductively coupled plasma mass spectrometry – part 2: polyatomic ions, *Spectrochim. Acta Part B* 56 (2001) 2479-2495.
- [11] A. Montaser, ed. *Inductively Coupled Plasma Mass Spectrometry*, New York: Wiley, 1998, 539-540, 765-769.

**CHAPTER 2. HIGH RESOLUTION STUDIES OF THE ORIGINS OF
POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS
SPECTROMETRY**

**PART I. IDENTIFICATION METHODS AND EFFECTS OF NEUTRAL GAS
DENSITY ASSUMPTIONS, EXTRACTION VOLTAGE, AND CONE MATERIAL**

A paper accepted for publication in *Spectrochimica Acta Part B* pending revisions

Jill Wisnewski Ferguson and R. S. Houk

Abstract

Common polyatomic ions (ArO^+ , NO^+ , H_2O^+ , H_3O^+ , Ar_2^+ , ArN^+ , OH^+ , ArH^+ , O_2^+) in inductively coupled plasma – mass spectrometry (ICP-MS) are identified using high mass resolution and studied using gas kinetic temperatures (T_{gas}) determined from a dissociation reaction approach. Methods for making accurate mass measurements, confirming ion identifications, and correcting for mass bias are discussed. The effects of sampler and skimmer cone composition and extraction voltage on polyatomic ion formation are also explored. Neutral species densities at several locations in the extraction interface are estimated and the corresponding effects of the T_{gas} value are calculated. The results provide information about the origins of background ions and indicate possible locations for their formation or removal.

Introduction

Inductively coupled plasma – mass spectrometry (ICP-MS) is a highly sensitive and selective technique for the identification and quantification of elements in diverse samples. Ideally, the ICP would produce only atomic ions. Unfortunately, some polyatomic ions

persist in the spectrum, despite the high temperature of the ICP, as noted in the early experiments [1]. The location(s) where these ions are made is a common, simple question with a complex answer. This set of papers continues previous work on this key issue [2, 3]. In particular, high mass resolution and accurate mass measurements are used to identify the polyatomic ions definitively, and ways to critically evaluate the conclusions about ion origins are discussed. The present method evaluates several experimental issues concerning use of magnetic sector instruments for such fundamental studies.

Conceivably, methods to discern the origins of undesirable polyatomic ions could lead to new methodology to suppress them, in addition to existing options such as high resolution [4, 5, 6, 7], cool plasma [8, 9, 10, 11], solvent removal [12, 13], and collision cells [14, 15, 16, 17]. The value of all these polyatomic ion attenuation methods for practical scientific applications would be enhanced if levels of polyatomic ions were lower initially.

Previous investigations

The origins of polyatomic ions in ICP-MS have been studied for some time by workers such as Nonose [9, 18, 19], Tanner [11, 17, 20], Becker [21], Evans [22], Beauchemin [23, 24], and Longerich [7, 25]. Most of this work has been done using quadrupole mass analyzers. In terms of high-resolution studies of polyatomic ions, Reed et al. [26] catalogued common interferences using accurate mass measurements, and Taylor et al. [7] identified and proposed structures for a number of unexpectedly abundant ions from fragments of carbohydrate molecules.

Theory

Overview

Our general method for investigating the origin of a particular polyatomic ion has been described [3]. Basically, a dissociation reaction is written involving two measurable ions A^+ and AB^+

$$K_d = \frac{n_{A^+}}{n_{AB^+}} n_B = f(T_{\text{gas}}, D_0) \quad (2)$$

where D_0 = dissociation energy of AB^+ and T_{gas} = gas kinetic temperature of region where AB^+ is made. A^+ may have two or more atoms; AB^+ has one more atom than A^+ .

The dissociation constant K_d is measured by determining the signal ratio for the pair of ions A^+/AB^+ , including corrections for mass bias. The density of the neutral product B is estimated from general knowledge of ICP conditions, e.g., solvent load. The value of n_B is itself a weak function of T_{gas} , so n_B is calculated iteratively for all temperatures. New refinements to this procedure that account for variable values of n_B , depending on the proposed location where the ion AB^+ is made, are discussed below.

The value of T_{gas} that yields the experimentally measured K_d value is evaluated by calculating the partition functions for the theoretical dissociation reaction

$$K_d = \frac{Z^{A^+} Z^B}{Z^{AB^+}} e^{\left[-D_0 / kT_{\text{gas}} \right]} = \frac{Z_{\text{trans}}^{A^+} Z_{\text{elec}}^{A^+} Z_{\text{trans}}^B Z_{\text{elec}}^B}{Z_{\text{trans}}^{AB^+} Z_{\text{elec}}^{AB^+} Z_{\text{vib}}^{AB^+} Z_{\text{rot}}^{AB^+}} e^{\left[-D_0 / kT_{\text{gas}} \right]} \quad (3)$$

Here Z = total partition function and z = partition function for translational, vibrational, rotational, and electronic degrees of freedom. Each partition function is also a function of temperature, so the theoretical K_d is calculated iteratively at each possible T_{gas} value and then compared with the measured K_d . Various reduced versions of Eq. 3 are available to simplify this task [3, 27].

Diagnosis of the origin of the polyatomic ion AB^+

The comparison of the theoretical and experimental dissociation equilibrium expressions yields a value of T_{gas} that is believed to be characteristic of the region where the ion is produced. Various ions are measured at abundances corresponding to different T_{gas} values, which is yet another indication that the spectra measured from the ICP do not reflect a source fully at equilibrium. Three general situations are common. First, some polyatomic ions are observed at relative abundances close to those expected from the ICP itself just outside the sampler cone. Here T_{gas} is typically 5500 K to 6000 K for an ICP operated under conditions that maximize atomic ion signals from injected analyte element M^+ . In our experience, strongly bound metal oxide ions like CeO^+ typify this situation. Our diagnosis is that this ion is present in the ICP, and its abundance is not altered greatly relative to that for Ce^+ by additional processes that might occur during extraction.

The second scenario is when the T_{gas} value that corresponds to the measured K_d value is lower than this cited range, 5500 – 6000 K. This means there is more polyatomic ion AB^+ in the spectrum than expected from the ICP itself. One explanation is that the dissociation reactions did not reach equilibrium in the ICP, as is apparently the case for the $C_xH_yO_z^+$ ions observed by Taylor and others [7, 28]. Alternatively, additional AB^+ ions could be made during the ion extraction process. ArO^+ is an example of an ion in this category.

The final situation is when the T_{gas} value that corresponds to the measured K_d value is higher than the range 5500 – 6000 K. This means there is less polyatomic ion AB^+ in the spectrum than expected from the ICP alone. Some process may occur that removes the polyatomic ion as it is extracted. NO^+ is one ion that displays gas kinetic temperatures in this regime. This correlates well with previous observations by Houk, who found less NO^+ than expected in the mass spectrum [3].

Experimental

Instrumental conditions

The instrument parameters were optimized to provide maximum signal for elemental standards. Typically, these resulted in "hot" plasma conditions (Table 1). The ThermoFinnigan ELEMENT 1 used for the experiments is a high resolution, double focusing magnetic sector device. Most experiments were conducted in medium resolution ($m/\Delta m \approx 4000$), which allowed for visualization of less intense ions. The experiments were often repeated in high resolution ($m/\Delta m \approx 12\,000$) to unequivocally identify unknown ions.

Low flow nebulizers were used in this experiment, as they are especially advantageous when introducing isotopically enriched or organic solvents. Solvent load to the plasma was measured to compare different nebulizers and for use in the equilibrium reactions to follow. Nebulizer uptake and drain output were measured repeatedly after solvent vapor in the spray chamber was allowed to equilibrate for about an hour. With a 100 $\mu\text{L}/\text{min}$ nebulizer, solvent loads for solutions of 1% nitric acid or deionized water were typically about 41 $\mu\text{L}/\text{min}$. A 20 $\mu\text{L}/\text{min}$ nebulizer produced a solvent load of about 23 $\mu\text{L}/\text{min}$.

Mass bias corrections

The experimentally-measured dissociation equilibrium expression comes from the measured signal ratio A^+/AB^+ . This signal ratio is vulnerable to mass bias, as are other ratio measurements in ICP-MS. The method for mass bias correction used here is based on the work of Sharp et al. [29]. The mass bias can be substantial because of the large differences in mass of A^+ and AB^+ , as much as 40 Da for Ar^+/Ar_2^+ and O^+/ArO^+ . Fortunately, the need for accurate corrections is much less extreme than is the case in isotope ratio measurements.

A typical mass bias plot is shown in Fig. 1. A single standard solution is prepared containing various elements at known concentrations, typically 10 ppb. Cerium was not used for mass bias measurement, since it is measured in subsequent experiments. The measured signals are adjusted for molar concentration, isotope abundance, and ionization efficiency at 7500 K and are plotted as shown in Fig. 1. Second order polynomial equations are fit to the plots for use in correcting measured signal ratios.

The full plot is used for high mass ions. For ions lighter than Ar_2^+ , a plot only up to $m/z = 103$ is utilized. This practice results in more stable and accurate mass bias corrections. For example, the mass bias between masses 28 and 29 is known to be 5.989% during a particular run, due to the direct measurement of silicon isotopes. Using the full mass bias plot generates a bias of 10.122%. However, using the plot only for lighter mass ions, a value of 5.066% is derived.

Repetition of mass bias measurements on different days always yields plots that look like those shown in Fig. 1. The exact values of the polynomial coefficients vary from day to day, so mass bias is measured at the beginning of each day's experiments.

Results and Discussion

Identification of ions by accurate mass measurements

The ThermoFinnigan ELEMENT is made to separate atomic analyte ions from interferences, not necessarily to measure their m/z values accurately. The ability of the instrument to do so is the subject of this section.

A typical spectrum at $m/z = 28$ is shown in Fig. 2. At low m/z values, the lightest ion observed is often an atomic ion, $^{28}\text{Si}^+$ in this case. These atomic ions serve as convenient mass calibration markers. Four peaks are easily resolved at medium resolution at $m/z = 28$. Low resolution measurements which assumed all of the signal at $m/z = 28$ was $^{28}\text{Si}^+$ would be incorrect. The peak assignment for the atomic ion can be easily verified by analyzing isotope ratios or an additional sample with a higher concentration of the element of interest.

The measured m/z value of the peak centroid (called mass hereafter for brevity) for the atomic ion is not exactly equal to the expected mass, due to imperfect mass calibrations, the mass of the electron(s) lost, and instrumental imprecision. The difference in these values is called mass offset and is very important in identification of the elemental formula of the polyatomic ion by accurate mass measurements. The mass offset varies with the magnetic field value, commonly called the magnet mass setting, as indicated in Table 2. The first four entries were obtained with the magnet mass set to $m/z = 59.930$. Note that the mass offset increases with distance from the magnet mass setting. When the magnet mass is changed to 68.925, a different progression of mass offset values develops. These mass offset gradients can guide the researcher in approximating the actual m/z of a peak. Therefore, it is important to observe an atomic ion (for determination of mass offset) and the polyatomic ions of interest in a single test at the same magnet mass setting.

Identification of polyatomic ions by isotope ratios

Isotope peaks and isotopically enriched solvents were used by Van Heuzen and Nibbering to verify identities of background ions in low-resolution measurements some time ago [30]. It is desirable to validate the initial ion identification from accurate mass measurements by examining isotope peaks.

Fig. 3 shows the isotopic verification of peaks identified as $^{40}\text{Ar}^{12}\text{C}^{16}\text{O}^+$ and $^{40}\text{Ar}^{13}\text{C}^{16}\text{O}^+$. $^{40}\text{Ar}^{13}\text{C}^{16}\text{O}^+$ (Fig. 3b) is observed at an intensity 1.09% of that for $^{40}\text{Ar}^{12}\text{C}^{16}\text{O}^+$ (Fig. 3a), compared to 1.12% predicted by the natural isotopic abundances. This small difference can be attributed to error in the measurement of the low signal intensity for the ^{13}C isotope. Therefore, the peak identities assigned from accurate m/z measurements are further validated.

In Fig. 4 the identification of $^{15}\text{N}^{16}\text{O}^+$ is confirmed via isotope ratio comparison with $^{14}\text{N}^{16}\text{O}^+$. The peak ascribed to $^{15}\text{N}^{16}\text{O}^+$ is present at 0.356% of $^{14}\text{N}^{16}\text{O}^+$ (data not shown), close to the value of 0.367% expected for the ^{15}N isotope. The tentative identification of $^{14}\text{N}^{17}\text{O}^+$ (the minor peak) is substantiated as well, because the intensity of this peak is too great to be $^{13}\text{C}^{18}\text{O}^+$, as confirmed by comparing to $^{13}\text{C}^{16}\text{O}^+$, which displays an integrated intensity of 1244 in another segment of this scan. This high resolution spectrum allows for separation of $^{14}\text{N}^{17}\text{O}^+$ from $^{14}\text{N}^{16}\text{OH}^+$, a much more intense peak. $^{14}\text{N}^{17}\text{O}^+$ is effectively hidden in a medium resolution mass spectrum.

D_2O (heavy water) can be used to shift peaks containing H atoms to confirm identifications [30, 31, 32]. Isopropanol (IPA) is a convenient additive to enhance the abundance of ions containing carbon for purposes of identification. On our particular instrument, the low-flow nebulizers used in this work allow introduction of up to 20% IPA without additional solvent removal procedures. Fig. 5 illustrates these capabilities. The

integrated intensities of $^{12}\text{C}_2^+$ and $^{12}\text{C}^{13}\text{C}^+$ seen in Fig. 5a and b are isotopically accurate to within 1.2%, sufficient to confirm the identity of both species initially suggested by accurate mass measurements. The identity of $^{12}\text{C}_2\text{H}^+$ (Fig. 5b) is confirmed by the observation of $^{12}\text{C}_2\text{D}^+$ from an IPA in D_2O solution (Fig. 5c). The intensities of these two peaks are similar, and neither exists in measurable amounts in nitric acid blanks using high resolution. In Fig. 5d, the major peak is $^{12}\text{C}_2^1\text{H}_2^+$ from 10% IPA in H_2O . This peak becomes $^{12}\text{C}_2\text{D}^+$ when the solvent is 10% IPA in D_2O , as seen in Figure 5c. Note that the $^{12}\text{C}_2\text{D}^+$ spectrum from the IPA/ D_2O solution is different from the $^{12}\text{C}_2\text{H}_2^+$ spectrum generated from IPA in deionized water. These two species are not resolvable by this instrument, but it is logical that the deuterated species is being formed. This is because it is assumed that D_2O would behave much the same way as H_2O , meaning it dissociates extensively and the resulting atoms are available for ionization and polyatomic ion formation (although they are not depleted substantially by these processes) [2, 3].

Comparison of T_{gas} values, magnetic sector vs. quadrupole instrument

Table 3 compares T_{gas} values measured in the present work to previous temperatures measured using a home-made quadrupole instrument in our lab. Both sets of measurements were made with a shielded load coil and hot plasma conditions. Both experiments used a nickel sampler cone, but the quadrupole device was run with a stainless steel skimmer, while the high resolution device used a nickel skimmer. The T_{gas} values measured for several sets of ions (ArN^+/N^+ , OH^+/O^+ , CeO^+/Ce^+) using the two different instruments are remarkably similar. There is only one ion for which the T_{gas} values are strikingly different; Ar_2^+ is apparently much less abundant relative to Ar^+ for the magnetic sector instrument. However, it is important to note that this T_{gas} value is very sensitive to the ion ratio measured. A very

small change in the measured ratio (≈ 0.0001) is significant (because of the high abundance of Ar^+ to Ar_2^+), and changes the calculated T_{gas} by about 1000 K.

One obvious difference between the two devices is the fact that in the sector instrument, the ions are quickly accelerated to about -2 kV right behind the skimmer cone. In contrast, the extraction lens was set to -70 volts in the quadrupole instrument. It makes sense that the commercial instrument with the greater accelerating voltage provides a calculated T_{gas} indicative of the polyatomic ion Ar_2^+ being removed, possibly through collisionally induced dissociation. Meanwhile, the homebuilt instrument with the lower accelerating voltage does not fragment the Ar_2^+ species, and a T_{gas} indicating formation of the polyatomic ion is observed. Nevertheless, the similarity of T_{gas} values for a given pair of ions between the two instruments argues that most ions have similar origins in the two devices, with the possible exception of Ar_2^+ . The T_{gas} values for the background ions are all much higher than the terminal T_{gas} of the supersonic expansion (~ 150 K), indicating the composition of the ion beam has “frozen” earlier in the extraction process. A similar effect occurs with ionization temperatures, which are typically ~ 7000 K, as expected in the ICP itself [33].

Effect of cone material

One curiosity evinced by Table 3 is the high value of T_{gas} for NO^+/N^+ . This effect is seen on both the quadrupole and magnetic sector instruments; the value is particularly high for the sector instrument. Perhaps there is a chemical reaction that occurs inside the sampler and/or skimmer cone tips that removes much of the NO^+ , resulting in a high T_{gas} value. NO^+ is isoelectronic with CO, which is known to be reactive on transition metal surfaces such as Ni and Pt, which are common cone materials [34]. By comparing the spectra generated by

cones of different materials, one may be able to investigate whether species are formed or removed chemically from the hot metal surface of sampler or skimmer cones.

To test this hypothesis, we used two types of metal cones on the sector instrument. Nickel cones were tested as well as cones made of aluminum, which is not a transition metal. Several sets of each type of cone were tested. This was done to ensure that the results seen were due to the differences in cone material, and not a random defect on one set of cones. The cones are supposed to have the same dimensions and sampler-skimmer tip spacing, although two different sets seldom yield exactly the same performance. The cones tend to have slightly different optimum operating conditions, requiring slightly different values of aerosol gas flow rate and RF power to attain maximum signal. This might indicate slight differences in thermal transfer from the plasma to the cones [35].

Aluminum cones generally have lower sensitivity for atomic ions in a multielement standard. This can be seen in Fig. 6a, which shows results taken on the same day with optimal conditions for four sets of cones, two nickel and two aluminum. The lower sensitivity for aluminum cones is not constant, as the sensitivity difference between the two types of cones changes across the mass range. This can be seen in Fig. 6b, which shows data collected from nickel cones for a 10 ppb standard, normalized to that gathered by aluminum cones. Interestingly, low mass ions are suppressed almost seven times more by aluminum cones than high mass ions. Note that this is separate from traditional mass bias, as elements at the same m/z are being compared for the two cone types. The data from plots such as Fig. 6b can be used to create a correction factor for this extra suppression related to mass to charge ratio. These differences must be accounted for if any comparisons are to be valid.

Fig. 7 shows the suppression factor estimated from Fig. 6b applied to background ion intensities. Ions seem to fall in three groups. The first is where the ion signals from

aluminum cones matches what is predicted from the sensitivity curve (Fig. 6b), giving a y axis value of approximately one. Ce^+ is the only such ion in this study. Ce^+ is still less abundant in spectra obtained by aluminum cones compared to nickel ones, but its intensity is proportional to the additional suppression normally seen. This further validates the bias correction method used here, as the adjustment based on other atomic standards accurately accounts for the suppression of this additional atomic ion.

The second group of ions (NO^+ , O_2^+ , Ar^+ , Ar_2^+ , ArH^+ , ArO^+ , H_3O^+ , and H_2O^+) are much more abundant from aluminum cones than is expected from the sensitivity curve (Fig. 6b). Finally, N^+ , O^+ , ArN^+ , OH^+ , and CeO^+ are more abundant in aluminum cone spectra than predicted, but are not nearly as abundant as those ions in the second group. Note that these ions are more abundant from aluminum cones than predicted by the suppression factor, but their raw intensity is still below that seen with nickel cones.

The NO^+/N^+ ratio is much larger with aluminum cones, as can be seen in Table 4. In comparing ion intensities adjusted for the increased mass bias from aluminum cones, it is found that the NO^+ intensity is suppressed much less than predicted, while the N^+ amount more closely resembles the behavior of atomic standards. Therefore, the ratio is larger because there is more NO^+ . It is possible that NO^+ , like CO, is reactive on hot transition metal surfaces, such as nickel cones. NO^+ may react at or cling to nickel cones, removing it from the spectra. NO^+ does not show this reactivity when aluminum sampler and skimmer cones are used. Therefore the overproduction/lower suppression seen with aluminum cones may be due to the "free" NO^+ being extracted into the mass spectrometer, whereas some NO^+ is trapped, neutralized or dissociated on the reactive nickel cones.

Only two ions, H_3O^+ and H_2O^+ , are more intense in spectra obtained with aluminum cones. We suggest the following tentative explanation. Perhaps the radical ion H_2O^+ is more

reactive on hot nickel than H_3O^+ , with its stable octet electronic configuration. It would then follow that more H_2O^+ would be seen in aluminum cone spectra. More H_3O^+ would then also be observed, because its precursor H_2O^+ is more abundant.

All of the ion ratios (polyatomic to dissociated ion) are larger with aluminum sampler and skimmer cones than with nickel (Table 4). This makes the calculated kinetic gas temperatures smaller. However, these changes are not so significant as to change the equilibrium assignment proposed. T_{gas} is mainly determined by the energies and partition functions of the ions. For many sets of ions, T_{gas} is not strongly dependent on the observed ion ratio, which can change with cone material.

Effect of assumed neutral gas densities on diagnosis of ion origin

For each dissociation reaction (Eq. 1), the measured value of K_d depends on an assumed value for n_B , the number density of the neutral product species, usually Ar, H, or O atoms. So far, these n_B values have been calculated based on their expected abundances in the ICP. This was considered valid as an initial approximation. Suppose a polyatomic ion yields a value of T_{gas} greatly different from that of the ICP region sampled. The value of n_B corresponding to conditions in different regions of the ion extraction process would be expected to differ from that outside in the ICP. Does use of an adjusted value of n_B , one deemed characteristic of the region where this ion is thought to be made, change the diagnosis of the ion's origin?

We examined three such scenarios. First, consider the density of B for the collision-rich region just behind the sampler (Fig. 8). This zone is clearly seen by the scattered light just behind the sampler in photographic studies of the region between the sampler and the skimmer [36, 37, 38]. The density here is described by

$$n(z) = 0.161(D_0 / z)^2 n_0 \quad (4)$$

where $n(z)$ is the density at axial position z , D_0 is the orifice diameter (1.0 mm), and n_0 is the original calculated density of the species in the ICP region sampled [33]. The density falls off rapidly enough that most collisions occur within the first 1 mm or so behind the sampler. We took the position $z = 0.5$ mm. Here the T_{gas} should drop to 4300 K if the ICP is 6000 K [33].

The second zone modeled is just before the skimmer tip, $z = 8.0$ mm. If the expansion is clean and there is no shock wave outside the skimmer tip, this expression should be valid. Here the gas density is reduced by a factor of 256 from that in the $z=0.5$ mm region described above. The temperature at the skimmer tip in a clean expansion should be much lower than that of the ICP, only about 150 K [33].

In the third scenario, a shock wave is assumed to occur near the skimmer tip. The shock wave could be just before the tip or possibly inside the throat of the skimmer [2, 33]. We multiply the gas density from a clean expansion by a factor of 8/5, in accord with that found by electron density measurements by Niu and Houk [2]. It is hard to say what T_{gas} should be in such a shock wave. Measurements of T_{gas} in a Mach disk (not quite the same thing, as no skimmer was present) gave values of ~ 3000 K [39]. We therefore estimate T_{gas} in a shock wave to be 2000 to 3000 K. If the shock wave is hotter than this estimate, the resulting T_{gas} will be closer to that of the ICP.

Table 5 lists comparative values of T_{gas} , all resulting from the same experimentally-measured ion signal ratio. The neutral density n_B is varied, corresponding to equilibrium formation of ions at three different positions during the ion extraction process. Some general observations are possible from this table. A position farther back in the extraction has a lower neutral density, and the kinetic gas temperature calculated for that position is lower.

Since use of an adjusted n_B value can only decrease T_{gas} , observation of T_{gas} values above that of the ICP clearly indicates net loss of the corresponding polyatomic ion during extraction. Such is the case for NO^+ in our experiments.

Except for possibly ArO^+/O^+ , the T_{gas} values resulting from the measured ion ratios are not consistent with equilibration of ions in a clean supersonic expansion, i.e., T_{gas} is way too high, AB^+ is not abundant enough. Even for ArO^+ , the measured T_{gas} is about 500 K, consistent with formation of some additional ArO^+ in the expansion, but not enough to reflect equilibrium conditions there. It is interesting to compare this result with that of Becker [21], who measured the abundance of ArM^+ ions, where the M^+ ions were from injected solutions, not the cone material. These ions are also weakly bound (D_0 values of 0.15 to 0.5 eV) and have neutral Ar atoms as the second atom. She reports a Boltzmann plot with a slope corresponding to $T_{\text{gas}} = 540$ K, similar to that found for ArO^+/Ar^+ in the present work. The fact that the signals for these ArM^+ ions all fall on one line indicates they have a common origin.

In both the present and previous work [3], a number of ions are observed at abundances consistent with $T_{\text{gas}} \approx 4300$ K, close to that expected at the $M \approx 1$ position 0.5 mm behind the sampler. Ions in this temperature range in both studies include ArN^+/N^+ , $\text{H}_2\text{O}^+/\text{OH}^+$, and OH^+/O^+ . The T_{gas} values from these ions are fairly close together, suggesting that they arise from a region where T_{gas} approaches equilibrium. Most of the collisions experienced by the ions have already occurred by the time they have traveled to the $M=1$ position. Since collisions are necessary to make more polyatomic ions, it is reasonable that additional amounts of some ions would be made there. It would be interesting to vary parameters in some way to further test the possibility that additional polyatomic ions can be made between the ICP and the $M=1$ position.

These experimental measurements were done on different days than those reported in Table 3. Assuming the ions to be present as such in the ICP, the resulting T_{gas} values in these two tables are very reproducible. The signal ratios CeO^+/Ce^+ , O_2^+/O^+ , and ArH^+/Ar^+ are the ones that are most sensitive in these experiments, as they change the most with T_{gas} .

Comparative T_{gas} values for these species between Tables 3 and 5 are 5720 vs. 5680 for CeO^+/Ce^+ , 4960 vs. 4950 for O_2^+/O^+ , and 5050 vs. 5040 for ArH^+/Ar^+ . Our method for reproducing plasma conditions based on maximizing the M^+ signal is very consistent day to day and is preferable to reproducing external parameters like aerosol gas flow rate, power and sampling position.

Effect of extraction voltage on ion ratios and T_{gas} values

The high extraction voltage in a sector instrument could affect the diagnosis of ion origin, especially if ions are made in a shock wave inside the skimmer tip [33] or dissociated by collisions with the gas in the beam through the skimmer. All studies reported above were done with -2000 volts on the extraction lenses (Fig. 8), the value typically used for analysis with this device. In this section, the extraction voltage was varied in 100 volt increments from -2000 to -1000 volts. Two general protocols were used. First, only the extraction voltage was changed; all other ICP and ion optical parameters were kept constant. Second, these other parameters were re-adjusted to maximize atomic ion signal from standards at each new extraction lens voltage. This latter procedure mimics actual analytical practice. Mass bias curves were measured at each new extraction lens setting, although the mass bias did not change greatly across the voltage range investigated.

Naturally, the basic transmission of the instrument changes with extraction voltage. The extent of the change depends on whether other conditions are re-optimized at each

extraction lens voltage. Fig. 9 compares sensitivity at -2000 volts to that at -1000 volts for atomic analyte ions. Behavior between these extremes was consistent and is not shown for brevity. If the operating parameters are re-optimized, this sensitivity ratio is fairly uniform and less than a factor of two for all analyte ions. If the other parameters are not re-optimized, the sensitivity ratio is larger and varies more. This variation from element to element is due in part to the poor signal to noise ratio for low m/z and low abundance ions at -1000 volt extraction voltage. However, a greater variation is still seen for the constant parameters scenario when comparing data from -1500 volt and -2000 volt extractions, a situation where signal intensities are improved.

These variations of atomic ion signals are thought to reflect the basic transmission of the instrument and are used to normalize the polyatomic ion signals as extraction voltage is changed (Fig. 10). Basically, if the ordinate in Fig. 10 is close to unity, the polyatomic ion signal is affected by extraction voltage to about the same extent as the atomic analyte ion. This is the situation for all ions if other parameters are kept constant at the two extraction voltages.

If the other parameters are re-optimized (Fig. 10), there are substantial variations in behavior for different background and polyatomic ions. In most cases the ratios for these ions remain within a window of 0.75 to 1.75, with three exceptions. Signals for Ar^+ and Ar_2^+ are reduced over two-fold more than the corresponding atomic analyte ion going from -2000 to -1000 volts. Thus, a higher extraction voltage generates much more Ar^+ and Ar_2^+ in the spectrum relative to M^+ ions. It is interesting to note that this ICP-MS device yields small but measurable levels of multiply charged Ar ions, including Ar^{3+} and Ar^{4+} [40]. These highly energetic species cannot be in the ICP and are probably made between the skimmer tip and the extraction lens. An electrical discharge between the skimmer and extraction lens,

or the presence of high energy electrons that have passed through the skimmer and then been accelerated back toward the skimmer tip by the high negative voltage on the extraction lens, could cause such energetic ionization processes. Few or no such ions are seen from quadrupole instruments. This effect would also enhance the abundance of Ar^+ , apparently slightly more than Ar_2^+ .

CeO^+ exhibits a value of well under 1.0 in Fig. 10. This indicates that the spectral intensity of CeO^+ is more stable with changes in extraction voltage than that of atomic ions. This ion is not as affected by a 1000 Volt drop as the ions produced from atomic standards. This ion is thought to be formed in the plasma, and the T_{gas} calculated is indicative of this process. Fig. 10 supports this, as it indicates that CeO^+ is not produced or removed by changes in extraction voltage.

Conclusions

Polyatomic ions can be correctly identified by accurate mass measurements using a high resolution instrument. The diagnosis of ion origins are similar for quadrupole and magnetic sector instruments, except that Ar_2^+ is dissociated by the high extraction voltage of the magnetic sector device. Identities can be confirmed using minor isotopes and/or isotopically enriched solvents. Mass bias is accounted for using curves developed from atomic standards. A second mass bias effect is seen in the comparison of aluminum and nickel cones. The signal difference between aluminum and nickel cones is greater at low m/z . Therefore, any assessment of the effects of cone material on polyatomic ion origin must take into account this graduated difference.

The dissociation equilibrium method used here to calculate gas kinetic temperatures can be used to study polyatomic ions in ICP-MS. Ratios of background ions can change

based on the cones used (aluminum or nickel) and instrument conditions, such as extraction voltage. This information can be used to maximize sensitivity for analytes at the same nominal mass as background ions for quadrupole devices. The calculated gas kinetic temperature does not change enough with manipulation of experimental parameters to alter the proposed formation or reduction regions or processes for the polyatomic ions studied. The ability to estimate neutral density at specific points in the interface region is displayed and then applied to the proposed dissociation reactions. This paper sets the experimental stage for further work into the origins and behavior of polyatomic ions in ICP-MS.

Acknowledgements

Ames Laboratory is operated for the U.S. Department of Energy by Iowa State University under Contract No. W-7405-Eng-82. This research was supported by the National Science Foundation through the Institute for Physical Research and Technology at ISU. JWF was also supported by the Velmer A. and Mary K. Fassel Fellowship. The nebulizers used were provided by Elemental Scientific Inc. The ICP-MS device was obtained with funds provided by the U. S. Department of Energy, Office of Nuclear Nonproliferation (NA-22) and the Office of Basic Energy Sciences.

References

- [1] R. S. Houk, V. A. Fassel, G. D. Flesch, H. J. Svec, A. L. Gray, C. E. Taylor, Inductively coupled argon plasma as an ion source for mass spectrometric determination of trace elements, *Anal. Chem.* 52 (1980) 2283-2289.
- [2] H. Niu, R. S. Houk, Fundamental aspects of ion extraction in inductively coupled plasma – mass spectrometry, *Spectrochim. Acta Part B* 51 (1996) 779-815.

- [3] R. S. Houk, N. Praphairaksit, Dissociation of polyatomic ions in the inductively coupled plasma, *Spectrochim. Acta Part B* 56 (2001) 1069-1096.
- [4] N. Bradshaw, E. F. H. Hall, N. E. Sanderson, Inductively coupled plasma as an ion source for high-resolution mass spectrometry, *J. Anal. Atom. Spectrom.* 8 (1989) 801-803.
- [5] I. Feldmann, W. Tittes, N. Jakubowski, D. Stuewer, U. Giessmann, Performance characteristics of inductively coupled plasma mass spectrometry with high mass resolution, *J. Anal. Atom. Spectrom.* 9 (1994) 1007-1014.
- [6] W. Tittes, N. Jakubowski, D. Stüwer, G. Tölg, J. A. C. Broekaert, Reduction of some selected spectral interferences in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 9 (1994) 1015-1020.
- [7] V. F. Taylor, R. E. March, H. P. Longerich, C. J. Stadey, A mass spectrometric study of glucose, sucrose, and fructose using an inductively coupled plasma and electrospray ionization, *Int. J. Mass Spectrom.* 243 (2005) 71-84.
- [8] S. J. Jiang, R. S. Houk, M. A. Stevens, Alleviation of overlap interferences for determination of potassium isotope ratios by inductively coupled plasma mass spectrometry, *Anal. Chem.* 60 (1988) 1217-1221.
- [9] N. S. Nonose, N. Matsuda, N. Fudagawa, M. Kubota, Some characteristics of polyatomic ion spectra in inductively coupled plasma – mass spectrometry, *Spectrochim. Acta Part B* 49 (1994) 955-974.
- [10] K. Sakata, K. Kawabata, Reduction of fundamental polyatomic ions in inductively coupled plasma mass spectrometry, *Spectrochim. Acta Part B* 49 (1994) 1027-1038.
- [11] S. D. Tanner, Characterization of ionization and matrix suppression in inductively coupled 'cold' plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 10 (1995) 905-921.

- [12] L. C. Alves, D. R. Wiedering, R. S. Houk, Reduction of polyatomic ion interferences in inductively coupled plasma mass spectrometry by cryogenic desolvation, *Anal. Chem.* 64 (1992) 1164-1169.
- [13] M. G. Minnich, R. S. Houk, Comparison of cryogenic and membrane desolvation for attenuation of oxide, hydride and hydroxide ions and ions containing chlorine in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 13 (1998) 167-174.
- [14] J. T. Rowan, R. S. Houk, Attenuation of polyatomic ion interferences in inductively coupled plasma mass spectrometry by gas-phase collisions, *Appl. Spectrosc.* 43 (1989) 976-980.
- [15] D. J. Douglas, Some current perspectives on ICP-MS, *Can. J. Spectroscopy* 34 (1989) 38-49.
- [16] G. C. Eiden, C. J. Barinaga, D. W. Koppenaal, Selective removal of plasma matrix ions in plasma source mass spectrometry, *J. Anal. Atom. Spectrom.* 11 (1996) 317-322.
- [17] S. D. Tanner, V. I. Baranov, D. R. Bandura, Reaction cells and collision cells for ICP-MS: a tutorial review, *Spectrochim. Acta Part B* 57 (2002) 1361-1452.
- [18] N. Nonose, M. Kubota, Non-spectral and spectral interferences in inductively coupled plasma high resolution mass spectrometry Part I: optical characteristics, *J. Anal. Atom. Spectrom.* 16 (2001) 551-559.
- [19] N. Nonose, M. Kubota, Non-spectral and spectral interferences in inductively coupled plasma high resolution mass spectrometry Part II: comparison of interferences in quadrupole and high resolution, *J. Anal. Atom. Spectrom.* 16 (2001) 560-566.
- [20] S. D. Tanner, Plasma temperature from ion kinetic energies and implications for the source of diatomic oxide ions in inductively coupled plasma – mass spectrometry, *J. Anal. Atom. Spectrom.* 8 (1993) 891-897.

- [21] J. S. Becker, H.-J. Dietze, Investigations on cluster and molecular ion formation by plasma mass spectrometry, *Fresenius J. Anal. Chem.* 359 (1997) 338-345.
- [22] E. H. Evans, L. Ebdon, L. Rowley, Comparative study of the determination of equilibrium dissociation temperature in inductively coupled plasma – mass spectrometry, *Spectrochim. Acta Part B* 57 (2002) 741-754.
- [23] M. M. Fraser, D. Beauchemin, Effect of concomitant elements on the distribution of ions in inductively coupled plasma – mass spectrometry – Part 2: polyatomic ions, *Spectrochim. Acta Part B* 56 (2001) 2479-2495.
- [24] A. E. Holliday, D. Beauchemin, Spatial profiling of ion distributions in a nitrogen-argon plasma in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 18 (2003) 289-295.
- [25] Guenther, D.; Longerich, H. P.; Jackson, S. E.; Forsythe, L. Effect of sampler orifice diameter on dry plasma inductively coupled plasma mass spectrometry (ICP-MS) backgrounds, sensitivities, and limits of detection using laser ablation sample introduction, *Fresenius J. Anal. Chem.* 355 (1996) 771-773.
- [26] N. M. Reed, R. O. Cairns, R. C. Hutton, Y. Takaku, Characterization of polyatomic ion interferences in inductively coupled plasma – mass spectrometry using a high resolution mass spectrometer, *J. Anal. Atom. Spectrom.* 9 (1994) 881-896.
- [27] R. S. Houk, Erratum to “Dissociation of polyatomic ions in inductively coupled plasma-mass spectrometry”, *Spectrochim. Acta Part B* 61 (2006) 235–236.
- [28] R. S. Houk, V. A. Fassel, H. J. Svec, Mass spectra of polar organic compounds in aqueous solutions introduced into an inductively coupled plasma, *Org. Mass Spectrom.*, 17 (1982) 240-244.

- [29] C. P. Ingle, B. L. Sharp, M. S. A. Horstwood, R. R. Parrish, D. J. Lewis, Instrument response functions, mass bias and matrix effects in isotope ratio measurements and semi-quantitative analysis by single and multi-collector ICP-MS, *J. Anal. Atom. Spectrom.* 18 (2003) 219-229.
- [30] A. A. van Heuzen, N. M. M. Nibbering, Elemental composition and origin of (polyatomic) ions in inductively coupled plasma mass spectrometry disclosed by means of isotope exchange experiments. *Spectrochim. Acta Part B* 48 (1993) 1013-1021.
- [31] F. G. Smith, R. S. Houk, Alleviation of polyatomic ion interferences for determination of chlorine isotope ratios by inductively coupled plasma – mass spectrometry, *J. Am. Soc. Mass Spectrom.* 1 (1990) 284-287.
- [32] M. V. Zoriy, L. Halicz, M. E. Ketterer, C. Pickhardt, P. Ostapczuk, J. S. Becker, Reduction of UH^+ formation for $^{236}\text{U}/^{238}\text{U}$ isotope ratio measurements at ultratrace levels in double focusing sector field ICP-MS using D_2O as solvent, *J. Anal. Atom. Spectrom.* 19 (2004) 362-367.
- [33] D. J. Douglas, J. B. French, Gas dynamics of the inductively coupled plasma mass spectrometry interface, *J. Anal. Atom. Spectrom.* 3 (1988) 743-747.
- [34] J. Joyce, A. Martin-Rodero, F. Flores, P. J. Grout, N. H. March, Indirect interactions between CO molecules on transition-metal surfaces and the interpretation of thermal desorption experiments, *J. Phys. C: Solid State Phys.* 20 (1987) 3381-3389.
- [35] S. A. Lehn, K. A. Warner, M. Huang, G. M. Hieftje, Effect of an inductively coupled plasma mass spectrometry sampler interface on electron temperature, electron number density, gas-kinetic temperature and analyte emission intensity upstream in the plasma, *Spectrochim. Acta Part B* 57 (2002) 1739-1751.

- [36] A. L. Gray, Stabilization of an inductively coupled plasma for inductively coupled plasma mass spectrometry with a flared torch extension, *J. Anal. Atom. Spectrom.* 7 (1992) 1151-1153.
- [37] A. T. Ince, J. G. Williams, A. L. Gray, Noise in inductively coupled plasma mass spectrometry: some preliminary measurements, *J. Anal. Atom. Spectrom.* 8 (1993) 899-903.
- [38] A. L. Gray, R. S. Houk, J. G. Williams, Langmuir probe potential measurements in the plasma and their correlation with mass spectral characteristics in inductively coupled plasma mass spectrometry, *J. Anal. Atom. Spectrom.* 2 (1987) 13-20.
- [39] H.B. Lim, R.S. Houk, M.C. Edelson, K.P. Carney, Some fundamental characteristics of a reduced-pressure plasma extracted from an ICP, *J. Anal. Atom. Spectrom.* 4 (1989) 365-370.
- [40] L. Moens, N. Jakubowski, Double-focusing mass spectrometers in ICP-MS, *Anal. Chem.* (1998) 251A-256A.

Tables

Table 1: Experimental Parameters

Instrument: ThermoFinnigan ELEMENT 1

Resolution: Medium ($m/\Delta m \approx 4000$), except where otherwise stated

Nebulizer: Elemental Scientific Inc. PFA 100 $\mu\text{L}/\text{min}$ or 20 $\mu\text{L}/\text{min}$ as needed

Spray Chamber: Teflon Scott type double pass

Cones: Nickel sampler and skimmer used most commonly, others as indicated

Power: 1200 W, changed marginally to improve performance

Gas flow rates: Optimized to produce greatest signal for standards

Solutions: Standards in 1% nitric acid. Other solutions used as indicated

Torch was shielded with the shield grounded

Table 2: Importance of magnet mass. Mass offset is difference between actual mass and measured mass. Note a) variation of mass offset with mass at constant magnet mass, and b) change of mass offset with magnet mass.

Magnet Mass	Ion	Measured Mass	Mass Offset
59.930	$^{64}\text{Zn}^+$	63.9203	-0.00885
59.930	$^{65}\text{Cu}^+$	64.9184	-0.00939
59.930	$^{66}\text{Zn}^+$	65.9166	-0.00944
59.930	$^{67}\text{Zn}^+$	66.9177	-0.00943
59.930	$^{68}\text{Zn}^+$	67.9152	-0.00965
68.925	$^{69}\text{Ga}^+$	68.9247	-0.00088
68.925	$^{70}\text{Zn}^+$	69.9244	-0.00092

Table 3: Comparison of measured T_{gas} values from background ions and CeO^+ between quadrupole and magnetic sector instrument. Nickel cones and shielded torch were used with the sector instrument. A nickel sampler, stainless steel skimmer, and shielded torch were used with the quadrupole instrument.

Ions	D_0 (eV)	T_{gas} Values (K)	
		Quadrupole [Ref. 3]	Magnetic Sector
ArO^+/O^+	0.312	1810	2930
$\text{Ar}_2^+/\text{Ar}^+$	1.2	4050	7030
ArN^+/N^+	1.866	4220	4340
OH^+/O^+	5.09	4690	4700
$\text{H}_2\text{O}^+/\text{OH}^+$	5.43	3170	3870
$\text{H}_3\text{O}^+/\text{H}_2\text{O}^+$	6.508	4410	4950
ArH^+/Ar^+	6.16	4190	5040
O_2^+/O^+	6.663	4780	4960
NO^+/N^+	11.76	6440	7330
CeO^+/Ce^+	8.81	5800	5720

Table 4: Comparison of measured T_{gas} values and signal ratios from background ions and CeO^+/Ce^+ . Ion ratios are corrected for basic mass bias. Magnetic sector instrument, nickel or aluminum cones.

Ions	D_0 (eV)	T_{gas} Values (K)		Signal Ratios	
		Ni Cones	Al Cones	Ni Cones	Al Cones
ArO^+/O^+	0.312	2880	1410	0.000387	0.00596
$\text{Ar}_2^+/\text{Ar}^+$	1.2	7180	6080	0.000211	0.000317
ArN^+/N^+	1.866	4370	3830	0.00729	0.0123
OH^+/O^+	5.09	4710	4470	0.107	0.243
$\text{H}_2\text{O}^+/\text{OH}^+$	5.43	3960	3560	0.723	9.12
$\text{H}_3\text{O}^+/\text{H}_2\text{O}^+$	6.508	4960	4750	0.119	0.247
ArH^+/Ar^+	6.16	5050	4880	0.213	0.366
O_2^+/O^+	6.663	5070	4590	0.0399	0.201
NO^+/N^+	11.76	7390	6850	0.224	1.13
CeO^+/Ce^+	8.81	5790	5540	0.0467	0.115

Table 5: Comparison of measured T_{gas} values from background ions and CeO^+/Ce^+ for a magnetic sector instrument with nickel cones and using various values of gas number density corresponding to different locations in the ion extraction process. If the dissociation reactions all reached equilibrium at the various positions, the measured T_{gas} values would be 4300 K 0.5 mm behind the sampler, about 150 K at the skimmer tip, and 2000 to 3000 K if there were a shock wave at the skimmer tip.

Ions	Calculated T_{gas} (K) for n_B calculated at following positions			
	In Plasma	0.5 mm behind Sampler	Clean Expansion Skimmer Tip	Shock Wave at Skimmer Tip
ArO^+/O^+	2830	2270	580	620
$\text{Ar}_2^+/\text{Ar}^+$	6840	5800	1900	2020
ArN^+/N^+	4510	4160	2100	2200
OH^+/O^+	4710	4580	3260	3340
$\text{H}_2\text{O}^+/\text{OH}^+$	3960	3870	2920	2980
$\text{H}_3\text{O}^+/\text{H}_2\text{O}^+$	4990	4730	3510	3590
ArH^+/Ar^+	5040	4910	3610	3690
O_2^+/O^+	4950	4830	3630	3710
NO^+/N^+	7400	7250	5650	5760
CeO^+/Ce^+	5680	5580	4380	4460

Figures

Figure 1: Mass bias plot for 10 ppb standards in medium resolution. This plot is used for ions above 80 Da.

Figure 2: Medium resolution mass spectrum showing $^{28}\text{Si}^+$ and polyatomic peaks from a solution of 10 ppb silicon in 20% isopropyl alcohol. There are 4 peaks of interest. The vertical axis has been limited to show the less intense peaks.

Peak Centroid	Integrated Intensity	Assigned Identity
27.9626	1.17×10^6	$^{28}\text{Si}^+$
27.9806	4.73×10^8	$^{12}\text{C}^{16}\text{O}^+$
27.9920	5.79×10^5	$^{14}\text{N}_2^+$
28.0045	7.07×10^4	$^{12}\text{C}^{14}\text{N}^1\text{H}_2^+$

Figure 3a: Medium resolution spectrum of $^{40}\text{Ar}^{12}\text{C}^{16}\text{O}^+$ from a nebulized solution of 20% isopropyl alcohol in deionized water. Peak centroid = 67.9538. Integrated intensity = 3.73×10^4 .

Figure 3b: Medium resolution spectrum of $^{40}\text{Ar}^{13}\text{C}^{16}\text{O}^+$ from a nebulized solution of 20% isopropyl alcohol in deionized water. Peak centroid = 68.9570. Integrated intensity = 409.

Figure 4: High resolution spectrum of $^{15}\text{N}^{16}\text{O}^+$, $^{14}\text{N}^{17}\text{O}^+$, and $^{14}\text{N}^{16}\text{OH}^+$ from a nebulized solution of 1% nitric acid in deionized water. $^{31}\text{P}^+$ would be to the left of the window shown.

Peak Centroid	Integrated Intensity	Assigned Identity
30.9810	3.69×10^4	$^{15}\text{N}^{16}\text{O}^+$
30.9884	4.64×10^3	$^{14}\text{N}^{17}\text{O}^+$
30.9919	7.36×10^4	$^{14}\text{N}^{16}\text{OH}^+$

High Resolution Spectrum

Figure 5a: High resolution spectrum of $^{12}\text{C}_2^+$ from 10% IPA in deionized water. Peak centroid = 23.9871. Integrated intensity = 3.63×10^5 .

Figure 5b: High resolution spectrum of $^{12}\text{C}^{13}\text{C}^+$ (low mass) and $^{12}\text{C}_2\text{H}^+$ (high mass) from 10% IPA in deionized water

Peak Centroid	Integrated Intensity	Assigned Identity
24.9901	8.04×10^3	$^{12}\text{C}^{13}\text{C}^+$
24.9946	4.42×10^3	$^{12}\text{C}_2\text{H}^+$

Figure 5c: High resolution spectrum of $^{12}\text{C}_2\text{D}^+$ from 10% IPA in D_2O . The unresolved signal to the low mass side may be $^{12}\text{C}^{13}\text{CH}^+$. Peak centroid = 26.0010. Integrated intensity = 3.53×10^3 .

Figure 5d: High resolution spectrum of $^{12}\text{C}_2\text{H}_2^+$ from 10% IPA in deionized water. Peak centroid = 26.0021. Integrated intensity = 1.77×10^3 .

Figure 6a: Mass bias plot comparing sensitivity from aluminum and nickel cones. Data taken from a 10 ppb multielement standard in medium resolution.

Figure 6b: Comparison of signal intensity from nickel cones versus aluminum cones.

Vertical axis is nickel data normalized to aluminum experiments. Data taken from a 10 ppb multielement standard in medium resolution.

Figure 7: Ion intensity comparison between nickel and aluminum cones normalized to the bias factor between nickel and aluminum cones.

Figure 8: Depiction of gas dynamics in the interface region.

Figure 9: Medium resolution data taken from 10 ppb multielement standard. Vertical axis shows response (mass bias) data at an extraction voltage of -2000 V normalized to that at -1000 V.

Figure 10: Comparison of ion intensities at -2000 V versus -1000 V. Ions of interest are normalized to ions from atomic standards.

**CHAPTER 3. HIGH RESOLUTION STUDIES OF THE ORIGINS OF
POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS
SPECTROMETRY**

PART II. NOVEL CALCULATIONS OF UNUSUAL IONS

A paper submitted to *Spectrochimica Acta Part B*

Jill Wisniewski Ferguson, Timothy J. Dudley, Kyle C. Sears, Mark S. Gordon,
and R. S. Houk

Abstract

Several polyatomic ions in ICP-MS are studied experimentally and by computational chemistry. Novel calculations based on spin-restricted open shell second order perturbation theory (ZAPT2) and coupled cluster (CCSD(T)) theory are performed to determine the energies and partition functions of the ions. These values are combined with experimental data to evaluate a dissociation constant for the proposed reaction. The outcome of this equation is a gas kinetic temperature (T_{gas}) value that describes the location where the polyatomic ion of interest is generated. The examination of N_2H^+ leads to a calculated T_{gas} of 4550 to 4900 K, depending on the theoretical modeling system used. The study of HCO^+ produces a similar temperature, which is not surprising considering the similar energies of HCO^+ and N_2H^+ . The evaluation of H_2CO^+ leads to a much lower T_{gas} (<1000 to 2000 K). Finally, the dissociation of H_3CO^+ generates a T_{gas} value between those from the other H_xCO^+ ions studied here. All of these calculated T_{gas} values correspond to formation of extra polyatomic ion in the interface or extraction region. The calculations reveal the existence of isomers such as HCO^+ and COH^+ , and H_2CO^+ and HCOH^+ , which complicate the interpretation of results.

Introduction

Inductively coupled plasma-mass spectrometry (ICP-MS) is a highly sensitive technique for the identification and quantification of elements in diverse samples. One strength of this technique is the ability of the ICP to atomize and ionize the sample. However, because of both the ICP's unique characteristics (high temperature, high concentration of electrons and argon) and the interface to the mass spectrometer, polyatomic ions may be created. Many of these ions do not readily exist in a normal environment, and little is known about them.

In this paper, unusual polyatomic ions are identified and quantified to some degree using a high resolution magnetic sector mass spectrometer (ThermoFinnigan ELEMENT). Calculations are done to determine the electronic, vibrational, and rotational energy states of the ions. With this information, both a theoretical and experimental dissociation expression can be written and compared. This comparison leads to the calculation of a T_{gas} value that describes the temperature of the region where the particular polyatomic ion is formed [1, 2]. Therefore, the origins of these previously unknown ions may be determined. Steps could then be taken to reduce the abundance of the polyatomic ions, thus improving the figures of merit of all ICP-MS devices.

Theory

In this work, the process of studying a polyatomic ion starts with writing a dissociation reaction for the ion of interest, as described elsewhere [1, 2, 3]. This dissociation reaction can be written as convenient for analytical measurement, as it does not need to take into account ionization or competing reactions. For a generic dissociation reaction such as

the experimental and theoretical dissociation equilibrium expressions are written and set equal (Eq. 2), and the gas kinetic temperature is evaluated

$$\frac{n_{A^+} n_B}{n_{AB^+}} = 10^{(1.5 \log T_{gas} \frac{5040 D_0}{T_{gas}} + 1.5 \log \frac{M_{A^+} M_B}{M_{AB^+}} + \log \frac{Z'_{A^+} Z'_B}{Z'_{AB^+}} + 20.274)} \quad (2)$$

where n is the number density of the indicated species (particles/cm³), D_0 is the dissociation energy of the ion AB^+ (eV), M is the atomic or molecular weight of the indicated species (g/mol), Z' is the product of the electronic, vibrational, and rotational partition function terms, and T_{gas} is the gas kinetic temperature (K). Note that the both the neutral species density and partition functions are themselves dependent on T_{gas} , and are calculated discretely for each T_{gas} . For a single atom species such as H or O, the Z' term only includes the electronic partition function [1].

The calculated T_{gas} identifies which way the system equilibrium lies and what process may influence the abundance of AB^+ . There are three major scenarios. See Table 1. The first is when the calculated T_{gas} is approximately 5500-6000 K. This indicates that the theoretical dissociation is roughly equal to the experimental dissociation ($K_{d, \text{theo}} \approx K_{d, \text{exp}}$). This finding suggests that the ions are formed in the ICP itself and is common for MO^+ ions such as CeO^+ .

The second arrangement is when the determined T_{gas} is less than 5500 K. The theoretical dissociation constant is larger than the experimental value ($K_{d, \text{theo}} > K_{d, \text{exp}}$). This suggests that extra AB^+ is being generated after the ICP, as the denominator of the experimental term is too large, making the experimental term smaller overall. The

polyatomic ion could be generated via collisions in the supersonic jet or in the extraction process. An example of this situation is the dissociation of H_2O^+ to OH^+ [2].

The third scenario is when the estimated T_{gas} is greater than 6000 K; the theoretical dissociation constant is smaller than the experimental value ($K_{\text{d, theo}} < K_{\text{d, exp}}$). Therefore AB^+ is being eliminated in the system, possibly through collision induced dissociation in the interface or extraction regions. NO^+ dissociating to N^+ models this behavior [1, 2].

Experimental

Table 2

In this work, some ions well known to the ICP-MS community are studied [1, 2] as well as others that have been scarcely discussed in this context (N_2H^+ , COH^+). Groups of ions of similar mass to charge ratio (m/z) are measured using a high resolution ICP-MS device, capable of resolution of approximately 300, 4000, or 11000. A single magnet mass setting is used for a particular group of ions to aid in their identification, as standard offsets develop between actual mass and measured mass at one magnetic field setting [2].

Once spectra are collected in a m/z range of interest, the offsets determined between the m/z values measured for M^+ ions from elemental standards and the accurate m/z values can be used to determine an approximate actual mass of the unknown ions. A line correction equation can be derived from standards to assist in actual mass estimation. The approximate actual m/z of unknown ions from the spectra are then compared to that of various combinations of atoms to help assign an identity. Suggested identities are confirmed by the examination of isotope peaks, addition of organic solvent, and/or nebulization of D_2O [2].

The vibrational, rotational, and electronic partition functions are then determined for the ions in question. Through our collaborative efforts, we attain energies calculated via the

CCSD(T) and ZAPT2 methods with a high degree of certainty. Articles by Gordon and coworkers explain the methods used and detail the work done on these specific ions [4, 5]. It should be noted that some of these ions have been studied previously. In those cases, the calculated values are consistent with previously published values [6-10]. Other ions have not been studied or modeled in this manner, so these computational results are of interest in their own right. Finally, a Microsoft Excel spreadsheet is set up to compare the experimental and theoretical dissociation constant expressions and ultimately calculate the gas kinetic temperature.

Results

N_2H^+ and N_2^+

N_2^+ is a common polyatomic ion seen in ICP-MS. The presence of N_2^+ in the mass spectrum is expected since the plasma flows into the air. However, the authors were surprised to note that N_2H^+ was more abundant than N_2^+ (Figs.1 and 2). Polyatomic ions with more atoms are generally expected to be less abundant at the high temperatures present in the ICP.

Once the true identity of the ions was clear via accurate m/z measurements, isotope ratio measurements, and solvent manipulation, we measured and modeled the following reaction.

The calculated energies for N_2H^+ are reported in Table 3.

A T_{gas} of about 4550 K is obtained using energies provided by the ZAPT2 calculation. Using the CCSD(T) data produces a kinetic gas temperature around 4900 K. The difference in temperature is primarily attributed to the variation in calculated

dissociation energy between the two models. The CCSD(T) method is the higher level calculation and thus is thought to be more accurate.

T_{gas} values of 4550 or 4900 K are lower than those seen from ions which originate in the plasma, such as CeO^+ . This indicates that additional N_2H^+ is being produced in the interface or extraction. It is reasonable that since there is more N_2H^+ than N_2^+ , it is being produced somewhere. In addition, N_2H^+ is more stable thermodynamically than N_2^+ . Once N_2H^+ is formed, it is more likely to remain than other reactive species, such as N_2 or N_2^+ . Therefore, it is not surprising to the computational chemist that N_2H^+ is more abundant than N_2^+ , and through this work, spectroscopists can begin to understand this initially surprising result.

Interestingly, when modeling the N_2H^+ system, it was found that Eq. 3 does not indicate the actual reaction mechanism. Instead, the following series of reactions is more likely to occur.

However, whether the calculations are done in one overall step (Eq. 3) or two steps (Eq. 4-5) using the principle,

$$K_{d_1} = K_{d_2} K_{d_3} \quad (6)$$

the same result for T_{gas} and probable ion origin is obtained. This speaks to the accuracy of the calculations and the application of the equilibrium principle to these polyatomic ions in ICP-MS.

HCO^+ , COH^+ , and CO^+

The dissociation of HCO^+ to $\text{CO}^+ + \text{H}$ is energetically similar to that of N_2H^+ and N_2^+ .

The reaction proceeds as

This system is similar to the one previously discussed in that HCO^+ is the most thermodynamically stable ion in Eq. 7. However, this situation is somewhat complicated by the fact that HCO^+ rearranges to COH^+ , and the two have different energies. The mass spectrometer cannot discriminate between these two species; the signal observed is that for the sum of HCO^+ and COH^+ . Since the overall energies of the ions and dissociation reactions are nearly the same for N_2H^+ and HCO^+ (Tables 3-5), a similar ion ratio and measured T_{gas} are expected. However, this is not precisely the case, as the $\text{HCO}^+/\text{COH}^+$ signal is less intense than that for CO^+ (Figs. 1-2). The $\text{N}_2\text{H}^+/\text{N}_2^+$ signal ratio is typically around 1.7, depending on conditions, while the signal ratio $(\text{HCO}^+/\text{COH}^+)/\text{CO}^+$ is about 0.4.

Eq. 2 is modified to account for the differing energies of COH^+ and HCO^+ which appear in the spectrum as aggregate COH^+ (not separate COH^+ and HCO^+).

$$\frac{n_{A^+} n_B}{n_{AB^+}} = 10^{(1.5 \log T_{\text{gas}} \frac{5040 D_0}{T_{\text{gas}}} + 1.5 \log \frac{M_{\text{CO}^+} M_H}{M_{\text{HCO}^+}} + \log \frac{Z'_{\text{CO}^+} Z'_H}{(Z'_{\text{HCO}^+} + Z'_{\text{COH}^+})} + 20.274)} \quad (8)$$

$$Z_{\text{elec COH}^+} = g e^{-D_{0p}/kT_{\text{gas}}} \quad (9)$$

D_0 in Equation 8 is the overall dissociation energy of the whole system, while D_{0p} in Equation 9 is the dissociation energy of COH^+ going to HCO^+ . g denotes the degeneracy of the electronic ground state of COH^+ .

Using Equation 8, the calculated kinetic gas temperature is about 4750 K using ZAPT2 theory or 4830 K using CCSD(T) theory. This means that HCO^+ (and probably COH^+) is being formed after the plasma. These temperature values are very similar to those calculated for the $\text{N}_2\text{H}^+/\text{N}_2^+$ system. This is the expected result based on the similar energies and partition functions for these groups of ions. However, this does not explain the very different intensities noted in the experimental spectra. One possible explanation is that there is more N_2 in the plasma than CO because of air infiltration in the argon gas lines, and this allows for a greater N_2H^+ production in the interface/extraction region in relation to that of HCO^+ . This would result in much larger N_2^+ and N_2H^+ peaks compared to those of HCO^+ and CO^+ , mirroring what is actually seen. Yet this still does not address the fact that the $\text{N}_2\text{H}^+/\text{N}_2^+$ ratio is much larger than that of HCO^+/CO^+ . Since the two reactions have similar energies and produce the same T_{gas} , the ion ratios are expected to be more similar than the results actually show. Why N_2H^+ is being formed seemingly preferentially to HCO^+ is unknown, and it may be related to the $\text{HCO}^+/\text{COH}^+$ rearrangement process. This relationship warrants further study.

H_2CO^+ and HCOH^+

In hoping to learn more about the peculiar behavior of HCO^+ and CO^+ , we studied H_2CO^+ as well. The H_2CO^+ system presented an additional challenge. Most of the ions studied here have their energies calculated by two methods, ZAPT2 and CCSD(T). The CCSD(T) method is the more sophisticated method, generally considered to be state-of-the-

art. In general, the ZAPT2 and CCSD(T) produce results similar enough that their differences make no significant difference in our final calculations. However, for the H_2CO^+ system, the two calculation methods point to different dissociation pathways. ZAPT2 theory (Eq. 10) indicates HCOH^+ is the lowest energy species, while CCSD(T) calculations (Eq. 11) point to HCOH^+ as the intermediate. Parallel calculations were done on each series, and the results were then compared.

Tables 6 and 7

When the standard matrix of 1% nitric acid is analyzed and calculations are done with ZAPT2 theory, a gas kinetic temperature of 850 K is calculated. The CCSD(T) method gives a T_{gas} of 1270 K. The introduction of 10-20% isopropyl alcohol, which increases the abundance of carbon containing species, results in a ZAPT2 determined temperature of 1360 K, and a CCSD(T) value of 2020 K. These low temperatures indicate formation of the larger polyatomic ions in the interface or extraction regions. The very low temperature assessed by the ZAPT2 method with 1% nitric acid is reasonable, as even lower temperatures (≈ 150 K) are estimated at the skimmer orifice [11]. The difference in calculated temperatures is not surprising, as the modeled reactions are different. Again, it makes sense that such a large ion would not survive the plasma, but instead be formed mainly in the interface or extraction.

The lower temperature calculated for H_2CO^+ , versus that for HCO^+ , seems to indicate that H_2CO^+ is formed farther back in the interface region, nearer to the skimmer cone. At this location in the zone of silence, the number density is 256 times lower than at a position closer to the sampler cone [3, 11]. Perhaps ion formation at this location is less probable because of the lower number density, in spite of the increased thermodynamic stability of the

larger polyatomic ion. This hypothesis on the balance between thermodynamics, gas dynamics, and ICP fundamentals, will be examined again later in this article.

H₂COH⁺

The dissociation reactions proposed for the final system studied in this series also differed slightly depending on the calculation level used. Eq. 12 is developed from ZAPT2 results, while Eq. 13 utilizes CCSD(T) theory.

H₂COH⁺ is the most thermodynamically stable ion studied here, but it is the least abundant ion in the mass spectrum. To observe H₂COH⁺, a 20% isopropyl alcohol solution was nebulized. Using the values in Tables 6 and 7, an average kinetic gas temperature of 3610 and 3650 K is calculated for the CCSD(T) and ZAPT2 systems respectively. This result is between the T_{gas} values determined for the HCO⁺ and HCOH⁺ systems.

Discussion

All of the gas kinetic temperatures calculated for the ions discussed in this article are below that of the ICP ($\approx$ 5500-6000 K). This means that additional amounts of these ions are created in the interface or possibly the extraction region through collisions with other species.

N₂H⁺ and HCO⁺ have similar structures and energies, and produce similar T_{gas} values, particularly when comparing the temperatures calculated from the more advanced CCSD(T) theory. The noted difference in ion ratios for the dissociation of N₂H⁺ and HCO⁺ may stem from the abundance of N₂ in the atmospheric pressure plasma, and could be influenced by the required rearrangement of HCO⁺ to COH⁺ for dissociation to occur.

The two hydrogen ions HCOH^+ and H_2CO^+ are more stable than $\text{HCO}^+/\text{COH}^+$, yet less abundant. The much lower T_{gas} points to formation farther back into the interface, where the density of species in the zone of silence is much lower. As the HCO^+ ion would have fewer opportunities to pick up one hydrogen atom at this cooler position, this would explain the lower intensity of $\text{HCOH}^+/\text{H}_2\text{CO}^+$ in the mass spectrum. The divergence in energies and dissociation mechanisms between the two computational theories also complicates this scenario.

H_2COH^+ is the most stable ion discussed here, but the least abundant in the mass spectrum. This is expected to some extent, since at the high temperatures of the ICP, polyatomic ions with more atoms would be less abundant. The calculations produce a T_{gas} in between the previous two values. This may mean that the ion is formed at a position in the zone of silence in between the other ions discussed here. It makes sense that H_2COH^+ , with three hydrogen atoms would have to be formed at a position of higher number density than $\text{HCOH}^+/\text{H}_2\text{CO}^+$, the two hydrogen counterparts.

However, the comparison between the two and one hydrogen species does not follow this logic, as HCO^+ is formed at a position closer to the sampler. Yet this reasoning may still hold true, as it is difficult to characterize where the required rearrangement of HCO^+ to COH^+ takes place, and if this location influences the temperature calculated for the dissociation to CO^+ . In addition, it is possible that the most thermodynamically stable form (HCO^+) is not what exists and is being detected at that precise mass to charge ratio. Perhaps COH^+ is the favored species in the interface, and the energy required to go from HCO^+ to COH^+ is provided by energetic collisions or carryover from the plasma itself.

Since our current instrumental technique offers no way to experimentally discriminate between rearranged species at the same m/z , we attempted to evaluate the abundance of each

isomer at specific T_{gas} values. For the isomers AB^+ and BA^+ a rearrangement reaction (Eq. 14) and corresponding equilibrium constant expression (Eq. 15) were written. This was then incorporated into equations (Eq. 16-17) which determined the amount of each isomer existing at a specific gas kinetic temperature, 5800 K, which describes the T_{gas} value of the plasma itself. Therefore, it is thought that these expressions estimate the proportion of signal to be attributed to each isomer originating in the plasma, if local equilibrium is reached.

$$K_{AB^+} = \frac{n_{BA^+}}{n_{AB^+}} = \frac{Z'_{BA^+}}{Z'_{AB^+}} e^{-D_0/kT_{\text{gas}}} \quad (15)$$

$$\alpha_{AB^+} = \frac{1}{1 + K_{AB^+}} = \frac{n_{AB^+}}{n_{AB^+} + n_{BA^+}} \quad (16)$$

$$n_{AB^+} = \alpha_{AB^+} N_t \quad (17)$$

All variables have been previously identified except N_t , which is the experimentally measured total amount of $(AB^+ + BA^+)$. These ion amounts (n) can then be incorporated into the previously described equations (Eq. 2), without requiring the summation of energies seen in Eq. 8.

In the comparison of isomers HCO^+ and COH^+ , it is found that at 5800 K, 53.1% of the signal at $m/z = 29.0027$ is attributed to the lower energy ion, HCO^+ . When comparing H_2CO^+ and $HCOH^+$ at 5800 K, 54.8% of the signal at $m/z = 30.0105$ is attributed to the lower energy species H_2CO^+ , according to the CCSD(T) theory. When the dissociation reactions were calculated with this adjustment, T_{gas} values were decreased by 150 – 200 K. This decrease does not substantially change the proposed location of ion formation in the interface

region. It is noteworthy that as α is calculated at higher temperatures, the proportion of the measured signal attributed to the higher energy ion increases. This result is logical, as the isomerization reactions are endothermic; as T_{gas} increases, the higher energy species becomes more prevalent. In addition, isomerizations that require more energy have a higher proportion of low energy species at a given temperature.

Conclusions

The dissociation of N_2H^+ to N_2^+ and that of HCO^+ to CO^+ are thermodynamically similar. This is evidenced by the similar kinetic gas temperatures calculated for each dissociation. Both N_2H^+ and HCO^+ appear to be formed in excess after the plasma, possibly in the interface region or during ion extraction. A similar process is happening in the creation of $\text{H}_2\text{CO}^+/\text{HCOH}^+$ and H_2COH^+ , which are more thermodynamically stable than HCO^+ or COH^+ . However, different T_{gas} values are calculated for these polyhydrogenated species, indicating that they may be formed in a different region or farther away from the plasma (closer to the terminal T_{gas} of the expansion). We need to discover what the balancing force is between thermodynamic stability, gas dynamics, and traditional ICP methodology. That may be the key in understanding the difference in ion ratios seen with the N_2H^+ to N_2^+ and HCO^+ to CO^+ systems.

Acknowledgements

This research was supported by the National Science Foundation through the Institute for Physical Research and Technology at ISU. JWF was also supported by the Velmer A. and Mary K. Fassel Fellowship. The nebulizers used were provided by Elemental Scientific Inc. The ICP-MS device was obtained with funds provided by the U. S. Department of

Energy, Office of Nuclear Nonproliferation (NA-22) and the Office of Basic Energy Sciences. Ames Laboratory is operated for the U.S. Department of Energy by Iowa State University under Contract No. W-7405-Eng-82.

References

- [1] R. S. Houk, N. Praphairaksit, Dissociation of polyatomic ions in the inductively coupled plasma, *Spectrochim. Acta Part B* 56 (2001) 1069-1096.
- [2] J. W. Ferguson, R. S. Houk, High resolution studies of the origins of polyatomic ions in inductively coupled plasma – mass spectrometry part I: identification methods and effects of neutral density assumptions, extraction voltage, and cone material, *Spectrochim. Acta Part B*, submitted May 2006.
- [3] H. Niu, R. S. Houk, Fundamental aspects of ion extraction in inductively coupled plasma mass spectrometry, *Spectrochim. Acta Part B* 51 (1996) 779-815.
- [4] M. W. Schmidt, K. K. Baldridge, J. A. Boatz, S. T. Elbert, M. S. Gordon, J. H. Jensen, S. Koseki, N. Matsunaga, K. A. Nguyen, S. Su, T. L. Windus, M. Dupuis, J. A. Montgomery, Jr., General atomic and molecular electronic structure system, *J. Comp. Chem.* 14 (1993) 1347-1363; M.S. Gordon and M.W. Schmidt, “Advances in Electronic Structure Theory: GAMESS a Decade Later”, *Theory and Applications of Computational Chemistry*, Ch.. 41, C. E. Dykstra, G. Frenking, K.S. Kim, G.E. Scuseria, Eds., Elsevier, 2005.
- [5] K. C. Sears, T. J. Dudley, J. W. Ferguson, R. S. Houk, M. S. Gordon, Theoretical investigations of polyatomic interferences in inductively coupled plasma-mass spectrometry experiments, Abstracts of Papers, 229th ACS National Meeting, San Diego, CA, United States, March 13-17, 2005.

- [6] J. Crovisier, Constants for molecules of astrophysical interest in the gas phase: photodissociation, microwave and infrared spectra. Version 4.2 (May 2002).
<http://www.usr.obspm.fr/~crovisie/basemole/>
- [7] NIST Chemistry Webbook, Vibrational and/or electronic energy levels.
<http://webbook.nist.gov>
- [8] J. Liu, B. Uselman, B. Van Devener, S. L. Anderson, Vibrational mode effects as a probe of inter-channel coupling in the reactions of formaldehyde cation with ammonia and water, J. Phys. Chem. A 108 (2004) 9945-9956.
- [9] J. Liu, B. Van Devener, S. L. Anderson, Collision-induced dissociation of formaldehyde cations: The effects of vibrational mode, collision energy, and impact parameter, J. Chem Phys. 116 (2002) 5530-5543.
- [10] P. Blowers, R. I. Masel, Calculated vibrational spectra for CH_nOH_m species, J. Phys. Chem A 104 (2000) 34-44.
- [11] D. J. Douglas, J. B. French, Gas dynamics of the inductively coupled plasma mass spectrometry interface,, J. Anal. Atom. Spectrom. 3 (1988) 743-747.
- [12] J. W. Ferguson, T. J. Dudley, R. S. Houk, A study of polyatomic ions in inductively coupled plasma – mass spectrometry Part III: CrO_xH_y^+ ions, in preparation.

Tables

Table 1: Examples of the three possible scenarios correlating T_{gas} with dissociation equilibrium

Equilibrium	Dissociation Reaction	T_{gas}	Diagnosis
$K_{\text{d, theo}} \approx K_{\text{d, exp}}$	$\text{CeO}^+ \rightleftharpoons \text{Ce}^+ + \text{O}$	5800 K	CeO^+ in ICP
$K_{\text{d, theo}} > K_{\text{d, exp}}$	$\text{H}_2\text{O}^+ \rightleftharpoons \text{OH}^+ + \text{H}$	3900 K	Extra H_2O^+ made in extraction
$K_{\text{d, theo}} < K_{\text{d, exp}}$	$\text{NO}^+ \rightleftharpoons \text{N}^+ + \text{O}$	7400 K	NO^+ lost during extraction

Table 2: Experimental conditions

Instrument: ThermoFinnigan ELEMENT 1

Resolution: Medium ($m/\Delta m \approx 4000$), except where otherwise stated

Nebulizer: Elemental Scientific Inc. PFA 100 $\mu\text{L}/\text{min}$ or 20 $\mu\text{L}/\text{min}$ as needed

Spray Chamber: Teflon Scott type double pass

Cones: Nickel sampler and skimmer

Power: 1200 W, changed marginally to improve performance

Gas flow rates: Optimized to produce greatest signal for standards

Solutions: Standards in 1% nitric acid. Other solutions used as indicated

Torch was shielded with the shield grounded

Table 3: Calculated properties of N_2H^+ by ZAPT2 and CCSD(T) theory.

	ZAPT2	CCSD(T)
Dissociation Energy (eV)	6.891	7.449
Vibrational Modes	4	4
Vibrational Energies (cm^{-1})	731.87	732.58
	731.87	732.58
	2151.1	2283.9
	3418.2	3434.9
Rotational Constant (cm^{-1})	1.536	1.557

Table 4: Calculated properties of HCO^+ , COH^+ , and CO^+ by ZAPT2 theory.

	HCO^+	COH^+	CO^+
Dissociation Energy to CO^+ (eV)	6.929	2.03	
Vibrational Modes	4	4	1
Vibrational Energies (cm^{-1})	859.76	321.80	2115.1
	859.76	321.80	
	2150.4	1761.9	
	3251.4	3349.4	
Rotational Constant (cm^{-1})	1.536	1.487	1.960

Table 5: Calculated properties of HCO^+ , COH^+ , and CO^+ by CCSD(T) theory.

	HCO^+	COH^+	CO^+
Dissociation Energy to CO^+ (eV)	6.925	1.75	
Vibrational Modes	4	4	1
Vibrational Energies (cm^{-1})	861.84	149.43	2203.1
	861.84	149.43	
	2212.3	1942.6	
	3327.1	3482.8	
Rotational Constant (cm^{-1})	1.487	1.487	1.939

Table 6: Calculated properties of H_2COH^+ , HCOH^+ , and H_2CO^+ by ZAPT2 theory.

	H_2COH^+	HCOH^+	H_2CO^+
Dissociation Energy to HCOH^+ (eV)	5.341	(0.845 to HCO^+)	0.13
Vibrational Modes	9	6	6
Vibrational Energies (cm^{-1})	1054.3	971.65	856.28
	1123.8	1001.5	1114.1
	1266.3	1248.9	1321.9
	1396.3	1727.8	1633.3
	1510.3	3122.0	2898.2
	1688.9	3497.4	3045.6
	3150.5		
	3301.4		
	3610.0		
Rotational Constant x (cm^{-1})	6.658	13.74	8.861
Rotational Constant y (cm^{-1})	1.153	1.271	1.348
Rotational Constant z (cm^{-1})	0.9869	1.167	1.174

Table 7: Calculated properties of H_2COH^+ , HCOH^+ , and H_2CO^+ by CCSD(T) theory.

	H_2COH^+	HCOH^+	H_2CO^+
Dissociation Energy to H_2CO^+ (eV)	5.17	0.23	(1.378 to HCO^+)
Vibrational Modes	9	6	6
Vibrational Energies (cm^{-1})	1038.3	969.57	855.58
	1121.0	1000.8	1071.0
	1246.8	1254.5	1272.6
	1400.4	1691.7	1663.2
	1488.7	3076.9	2811.4
	1663.2	3515.4	2920.5
	3109.5		
	3255.5		
	3620.4		
Rotational Constant x (cm^{-1})	6.637	13.68	8.917
Rotational Constant y (cm^{-1})	1.146	1.264	1.327
Rotational Constant z (cm^{-1})	0.9800	1.160	1.160

Table 8: Gas kinetic temperature results

Ion	Method	Solution	T _{gas} (K)
N ₂ H ⁺	ZAPT2	1% HNO ₃	4550
N ₂ H ⁺	CCSD(T)	1% HNO ₃	4900
HCO ⁺	ZAPT2	1% HNO ₃	4750
HCO ⁺	CCSD(T)	1% HNO ₃	4830
HCOH ⁺	ZAPT2	1% HNO ₃	850
HCOH ⁺	ZAPT2	10-20% IPA	1360
H ₂ CO ⁺	CCSD(T)	1% HNO ₃	1270
H ₂ CO ⁺	CCSD(T)	10-20% IPA	2020
H ₂ COH ⁺	ZAPT2	10-20% IPA	3650
H ₂ COH ⁺	CCSD(T)	10-20% IPA	3610

Figures

Figure 1: Medium resolution spectrum of CO^+ (left or low mass peak) and N_2^+ (right or high mass peak) in 1% nitric acid

Figure 2: Medium resolution spectrum of HCO^+ (left or low mass peak) and N_2H^+ (right or high mass peak) in 1% nitric acid

**CHAPTER 4. HIGH RESOLUTION STUDIES OF THE ORIGINS OF
POLYATOMIC IONS IN INDUCTIVELY COUPLED PLASMA – MASS
SPECTROMETRY**

PART III. CrO_xH_y^+ IONS

A paper to be submitted to *Spectrochimica Acta Part B*

Jill Wisnewski Ferguson, Timothy J. Dudley, Mark S. Gordon, and R. S. Houk

Abstract

Chromium polyatomic ions in ICP-MS are studied using a high resolution ICP-MS device. These ions include CrO^+ , CrOH^+ , CrO_2^+ , CrO_2H^+ , CrO_2H_2^+ , CrO_3^+ , CrO_3H^+ , and CrO_3H_2^+ . Accurate m/z measurements are made of these previously unidentified ions. Ion identities are confirmed using isotope ratios. The prevalence of these rather abundant CrO_xH_y^+ species is found to be independent of the oxidation state of chromium introduced into the ICP. Elements surrounding chromium on the periodic table, such as W, Mo, V, Ti, Mn, and Fe, are also studied. Results indicate that chromium is unique in its polyatomic ion formation. Finally, dissociation constants are evaluated to determine a gas kinetic temperature (T_{gas}) for the dissociation of CrOH^+ to CrO^+ that is thought to describe the location of ion formation in the ICP-MS device.

Introduction

Platinum group elements (PGEs), such as Pt, Pd, and Rh, are often used as catalysts for the control of automobile and industrial emissions. While this use limits the harmful byproducts of automobile exhausts, it also leads to PGE pollution, especially near roadways

[1, 2, 3]. The levels of these elements are often measured by ICP-MS. Chromium is also monitored often in environmental and clinical samples and has catalytic properties.

High resolution ICP-MS devices offer a distinct advantage in chromium measurements, as common polyatomic ions are present at $m/z = 52$, 53 , and 54 . However, it is often thought that quadrupole devices are sufficient for the detection of palladium, rhodium, and ruthenium, as common polyatomic ions are not prevalent at those m/z levels. Matrix separation, primarily employing ion exchange columns, has been used to diminish the appearance of ions such as $^{40}\text{Ar}^{65}\text{Cu}^+$, $^{89}\text{Y}^{16}\text{O}^+$, and $^{179}\text{Hf}^{16}\text{O}^+$ during the determination of platinum and palladium in environmental samples such as dust [4, 5, 6, 7]. However, results for Pd even after separation indicate recovery of over 200% in some instances, indicating the presence of unidentified interferences [7].

Researchers have not fully identified and compared polyatomic interferences stemming from chromium at these masses. Krushevska and co-workers documented interferences on rhodium seen in their laboratory with a high chromium matrix [8]. However, only the chromium interferences at m/z 102 and 103 were discussed. Polyatomic ion identities for these masses were proposed but not confirmed.

This paper describes remarkable chromium polyatomic ions such as CrO_2^+ , CrO_2H_2^+ , and CrO_3H^+ which obscure the measurement of Sr, Rb, Ru, Rh, and Pd. These ions are unequivocally identified by the use of chromium isotope ratios and accurate m/z measurements. While polyatomic ions with more atoms are generally expected to be less abundant from the ICP, this is not the case with these notable chromium species. The effect of the oxidation state of the chromium introduced into the ICP is also explored. Metals surrounding chromium on the periodic table, such as Ti, V, Mn, Fe, Mo, and W, are also studied to determine if chromium is unique in its polyatomic ion formation. Experimental

measurements and theoretical data are combined to evaluate a dissociation constant for polyatomic ions. This generates a T_{gas} value which points to the location of polyatomic ion formation.

Experimental

All experiments were done on a ThermoFinnigan ELEMENT 1 in medium or high resolution, $m/\Delta m \approx 4000$ and 11000 respectively. Solutions of 50 ppm Cr in 1% nitric acid were nebulized with a 100 $\mu\text{L}/\text{min}$ PFA nebulizer from Elemental Scientific Inc. This concentration was chosen to mimic industrial measurement of palladium, where some unknown and unusual polyatomics were first noticed with a quadrupole device [8]. Solutions of 50 ppm Ti, V, Mn, Fe, Mo, and W in 1% nitric acid were also analyzed as a basis for comparison. Nickel cones were used and conditions such as power and gas flows were optimized to maximize signal for atomic standards. Mass bias measurements were conducted as outlined by Ingle, Sharp, and coworkers [9], so that signals at different m/z could be accurately compared.

Polyatomic ion identification techniques have been previously discussed in more detail by the authors [10, 11, 12]. Accurate mass measurements are made using wide m/z acquisition windows, so known atomic ions can be captured as well. These atomic ions provide a reference m/z measurement that determines the difference between the actual and measured masses of ions at a particular magnet mass setting. Measurements are made in medium and high resolutions. Identifications proposed for the polyatomic ions are verified by isotope ratios.

Results and Discussion

Discovery and Identification of Ions

Unusually intense signals were repeatedly seen in the ranges $m/z = 83.92 - 87.95$ and $97.92 - 103.95$ when nebulizing solutions of 50 ppm Cr. These ions were not observed from blank solutions of 1% nitric acid, nor from solutions of Ru, Rh, or Rb. The signals from these unidentified ions were equal to or more intense than the M^+ count rate obtained from 10 ppb solutions of Rb, Ru, and Rh. The natural chromium isotope ratios were observed in these unknown peaks, and were eventually used in confirming the polyatomic ion identities. Spectra are shown in Figures 1 and 2.

The discovery of additional ion peaks in high resolution is an important progression in the unequivocal identifications put forth here. For example, medium resolution spectra of 50 ppm Cr at $m/z=86$ show two peaks which are not fully resolved at m/z 85.9315 and 85.9474 (Figure 1a). These peaks could be identified as $^{54}\text{Cr}^{16}\text{O}_2^+$ and $^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$. However, an examination of this same mass range in high resolution (Figure 1b) uncovers three distinct ions in this range. Signals at 85.9385 ($^{54}\text{Cr}^{16}\text{O}_2^+$), 85.9476 ($^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$), and 85.9555 ($^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$) are observed, and isotope ratio matches can be made for $^{54}\text{Cr}^{16}\text{O}_2^+$ and $^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$ with other chromium polyatomic ions. $^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$ cannot be confirmed via isotope ratios, because it is not fully resolved from $^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$, a more intense peak, in high resolution. However, mass offsets, as well as the isotope ratio confirmed identifications of $^{50}\text{Cr}^{16}\text{O}_2\text{H}^+$, $^{52}\text{Cr}^{16}\text{O}_2\text{H}^+$, and $^{54}\text{Cr}^{16}\text{O}_2\text{H}^+$ all support the peak at 85.9476 being identified as $^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$.

The ions listed in Table 1 all contain one chromium atom with oxygen and often hydrogen. Spectra were also obtained in the m/z range ascribed to polyatomic ions with more than one chromium atom. However, these ions were much less abundant than their single

chromium counterparts. In addition, the resolution required to resolve any probable ions is higher than that achievable from the high resolution device employed. In any case, it is known that for neutral Cr_2O_x molecules, it takes about 1 eV to break off a Cr atom, the lowest energy required of any possible dissociation. Removing one oxygen requires about 5 eV. Splitting the molecule into two, each containing a chromium atom, is the intermediate energy pathway. So it is likely that a Cr_2O_x molecule would lose a chromium atom, as it is the lowest energy dissociation pathway. Additionally, the Cr-O bonds are stronger in CrO_x than in Cr_2O_x molecules, as determined from the bond lengths calculated for these molecules [13]. This information seems to further support the high abundance of the CrO_xH_y^+ species relative to $\text{Cr}_2\text{O}_x\text{H}_y^+$.

Generally, the abundance of polyatomic ions in ICP-MS should decrease as the number of atoms in the polyatomic species increases. For instance, ArH^+ is more abundant than ArH_2^+ . Chromium polyatomic ions blatantly breach this commonly held principle. Figure 3 outlines some of the unusual ion signals seen. This figure accounts for mass bias, which traditionally makes ions of greater m/z more intense than ions of lower m/z at the same concentration. Species up to CrO_3H_2^+ are abundant. CrO_2^+ is more abundant than CrO^+ , CrO_3H^+ is more profuse than CrO_3^+ , and CrO_3H^+ is nearly as intense as CrO^+ . These results are surprising and lead to further investigation of chromium polyatomic ions in ICP-MS.

Chromium Oxidation State and Polyatomic Ions

While Cr(III) is not generally considered hazardous, Cr(VI) is a known carcinogen. Solutions of 50 ppm Cr(VI), 50 ppm Cr(III), and 50 ppm Cr of an unknown oxidation state were all tested separately to see if differences existed in the formation of these unusual polyatomic ions. Figure 4 compares the results from chromium solutions of different

oxidation states. Slight differences in individual solution concentrations are accounted for in this figure. The solution with the unknown oxidation state consistently produced less polyatomic ions than either the Cr(III) or Cr(VI) solutions. This could be due to a variety of factors, including the age or origin of the standard. However, it does not appear between the Cr(III) and Cr(VI) solutions that one steadily produces more polyatomic ions than the other. This result is important to note, since it confirms that all chromium solutions produce polyatomic interferences on Rb, Ru, Rh, and Pd, not just solutions that contain chromium in a specific oxidation state. This also indicates that the effect of the plasma, interface, and/or ion extraction is more influential in creating these polyatomic interferences than the original solution composition.

Comparison of Polyatomic Ions from Chromium and Other Metals

The polyatomic ions generated from chromium are unusually intense and include a large number of atoms. Next, the elements surrounding chromium on the periodic table were studied to determine if chromium is unique in this trait. Ti, V, Mn, Fe, Mo, and W were chosen for study because of their proximity to Cr on the periodic table. Solutions of 50 ppm of the metal of interest were individually introduced and m/z values corresponding to MO_xH_y^+ species, where $x=1-3$ and $y=0-3$, were monitored.

Figure 5 outlines the comparison between all the metals tested. Adjustments for mass bias, isotopic abundance, and concentration have been made. This includes a molar term, so elements of varying atomic weights can be accurately compared on a mole to mole basis. In addition, these data are normalized to the abundance of the analogous CrO_xH_y^+ ion from chromium of an unknown oxidation state, i.e. chromium has a consistent value of 1.0 on the y-axis for each ion shown.

W, Mo, Ti, and V all generate more of the MO^+ species than Cr does. This result is not surprising, as these metal cations all have higher oxygen atom affinities than the chromium cation. In an ICP/SIFT (selected ion flow tube) device the M^+ ions from these four elements all undergo a bimolecular oxygen atom transfer, whereas chromium exhibits a slow O_2 addition pathway, resulting in CrO_2^+ [14]. Similar results were seen in surface ionization ion beam experiments [15].

The only other ion that is more abundant than its chromium counterpart is MoO_3^+ . Kretzschmar et al. calculated the energies of MoO_x^+ , where $x=0-3$ [16]. The three oxygen species was found to have the lowest energy, making it the most stable ion. This may be why the abundance of MoO_3^+ is higher than expected, particularly in comparison with chromium. However, the balancing of thermodynamic stability and ICP methodology still needs to be explored.

Gas kinetic temperature

Experimental data were combined with calculated energies and partition functions of the ions to evaluate dissociation constants. The resulting T_{gas} value indicates the polyatomic ion formation or removal process [10, 11, 12]. The reaction studied was

The energies of states were specifically calculated for our purposes here and are reported in Table 2. These values are similar to those calculated for neutrals by Espelid [17]. The methodology of the calculations will be described separately [18].

A T_{gas} value of 850 K is calculated for the ratio $\text{CrOH}^+/\text{CrO}^+$ using only the ground state of each ion. If the lowest excited electronic state of CrOH^+ is included in the partition function, the temperature is decreased by 10 K. An alternative calculation was also employed, which estimates the amount of CrOH^+ in the excited state at 5800 K, which is the gas kinetic temperature of the region of the plasma sampled (determined by the measurement of CeO^+). The portion of the CrOH^+ signal attributed to the excited state then dissociates into CrO^+ . The T_{gas} value for this scenario is 600 K. These numbers are not remarkably different, and they all indicate that CrOH^+ is being formed near the terminus of the interface region. This result is encouraging, as it suggests that the abundance of these polyatomic ions may be reduced with changes in the design of the interface region.

Chromium dioxide cation was also studied. This ion is more abundant than either CrO^+ or CrOH^+ . The reaction studied was

Again, the energies of states were expressly calculated for our purposes here and are shown in Table 3 [18]. CrO_2^+ has two degenerate ground states. The energies calculated agree with the work published by Fiedler [19].

The ion ratio $\text{CrO}_2^+/\text{CrO}^+$ corresponds to a T_{gas} value of 2100 K, indicating excess formation of CrO_2^+ after the plasma. Including the low lying electronic states of CrO_2^+ and CrO^+ only lowers the temperature by 50 K. It appears that CrO_2^+ formation occurs at a region of higher gas density in the interface than that of CrOH^+ . CrO_2^+ is more abundant than CrOH^+ . There are an increased number of collisions possible at the 2100 K region of the interface compared to a position closer to the skimmer corresponding to 850 K. This may explain the relative abundances seen of these two ions.

Conclusions

Polyatomic ions in the form CrO_xH_y^+ are abundant and can obscure the measurement of Rb, Sr, Ru, Rh, and Pd, among other elements. Other metals near chromium on the periodic table were investigated, and chromium was the only metal tested that formed such abundant polyatomic ions. As the intensity of these polyatomic ions is not related to the oxidation state of chromium, it is important to determine the chromium levels in samples before determining the PGE concentration by quadrupole ICP-MS. By calculating the gas kinetic temperature of CrOH^+ and CrO_2^+ , it is determined that both of these ions are produced in excess after the plasma, most likely in the interface region. Further work is being done to calculate the T_{gas} values of more of these complicated ions, to gain insight into their formation or removal processes.

Acknowledgements

The authors wish to thank Antoaneta Krushevska for bringing her interesting palladium analysis problem to our attention. This research was supported by the National Science Foundation through the Institute for Physical Research and Technology at ISU. JWF was also supported by the Velmer A. and Mary K. Fassel Fellowship. The nebulizers used were provided by Elemental Scientific Inc. The ICP-MS device was obtained with funds provided by the U. S. Department of Energy, Office of Nuclear Nonproliferation (NA-22) and the Office of Basic Energy Sciences. Ames Laboratory is operated for the U.S. Department of Energy by Iowa State University under Contract No. W-7405-Eng-82.

References

- [1] E. Helmers, Platinum emission rate of automobiles with catalytic converters; comparison and assessment of results from various approaches, *Environ. Sci. Pollut. Res.* 4 (1997) 100-103.
- [2] J. Schafer, J. Echardt, Z. A. Berner, D. Stuben, Time-dependent increase of traffic-emitted platinum group elements (PGE) in different environmental compartments, *Environ. Sci. and Tech.* 33 (1999) 3166-3170.
- [3] K. Ravindra, L. Bencs, R. van Grieken, Platinum group elements in the environment and their health risk, *Sci. Tot. Environ.* 318 (2004) 1-43.
- [4] I. Jarvis, M. M. Totland, K. E. Jarvis, Assessment of dowex-1-X8-based anion exchange procedures for the separation and determination of Ru, Rh, Pd, Ir, Pt, and Au in geological samples by inductively coupled plasma – mass spectrometry, *Analyst* 122 (1997) 19-26.
- [5] I. Jarvis, M. M. Totland, K. E. Jarvis, Determination of the platinum group elements in geological materials by ICP-MS using microwave digestion, alkali fusion, and cation exchange chromatography, *Chem. Geol.* 143 (1997) 27-42.
- [6] J. D. Whiteley, F. Murray, Determination of platinum group elements (PGE) in environmental samples by ICP-MS: a critical assessment of matrix separation for the mitigation of interferences, *Geochem.: Explor., Environ., Anal.* 5 (2005) 3-10.
- [7] B. A. Lesniewska, B. Godlewska-Zylkiewicz, A. Ruszczynska, E. Bulska, A. Hulanicki, Elimination of interferences in determination of platinum and palladium in environmental samples by inductively coupled plasma – mass spectrometry, *Anal. Chim. Acta* 564 (2006) 236-242.

- [8] A. P. Krushevska, Y. Zhou, V. Ravikumar, Y.-J. Kim, J. Hinrichs, Chromium based polyatomic interferences on rhodium in ICP-MS, *Spectrochim. Acta Part B*, submitted 2006.
- [9] C. P. Ingle, B. L. Sharp, M. S. A. Horstwood, R. R. Parrish, D. J. Lewis, Instrument response functions, mass bias and matrix effects in isotope ratio measurements and semi-quantitative analysis by single and multi-collector ICP-MS, *J. Anal. Atom. Spectrom.* 18 (2003) 219-229.
- [10] R. S. Houk, N. Praphairaksit, Dissociation of polyatomic ions in the inductively coupled plasma, *Spectrochim. Acta Part B* 56 (2001) 1069-1096.
- [11] J. W. Ferguson, R. S. Houk, High resolution studies of the origins of polyatomic ions in inductively coupled plasma – mass spectrometry part 1: identification methods and effects of neutral gas density assumptions, extraction voltage, and cone material, *Spectrochim. Acta Part B*, submitted May 2006.
- [12] J. W. Ferguson, T. J. Dudley, K. C. Sears, M. S. Gordon, R. S. Houk, High resolution studies of the origins of polyatomic ions in inductively coupled plasma – mass spectrometry part 2: novel calculations of unusual ions, *Spectrochim. Acta Part B*, submitted June 2006.
- [13] S. Veliah, K. Xiang, R. Pandey, J. M. Recio, J. M. Newsam, Density functionality study of chromium oxide clusters: structures, bonding, vibrations, and stability, *J. Phys. Chem. B* 102 (1998) 1126-1135.
- [14] G. K. Koyanagi, D. Caraiman, V. Blagojevic, and D. K. Bohme, Gas-phase reactions of transition metal ions with molecular oxygen: room-temperature kinetics and periodicities in reactivity, *J. Phys. Chem. A* 106 (2002) 4581-4590.

- [15] E. R. Fisher, J. L. Elkind, D. E. Clemmer, R. Georgiadis, S. K. Loh, Aristov, N, L. S. Sunderlin, P. B. Armentrout, Reactions of fourth-period metal ions (calcium(1+)-zinc(1+)) with oxygen: metal-oxide ion bond energies, *J. Chem. Phys.* 93 (1990) 2676-2691.
- [16] I. Kretzschmar, A. Fieldler, J. N. Harvey, D. Schroder, and H. Schwartz, Effects of sequential ligation of molybdenum cation by chalcogenides on electronic structure and gas-phase reactivity, *J. Phys. Chem. A* 101 (1997) 6252-6264.
- [17] O. Espelid, K. J. Borge, V. R. Jensen, Structure and thermodynamics of gaseous oxides, hydroxides, and mixed oxohydroxides of chromium: $\text{CrO}_m(\text{OH})_n$ ($m, n = 0-2$) and CrO_3 . A computational study, *J. Phys Chem. A* 102 (1998) 10414-10423.
- [18] T. J. Dudley, M. S. Gordon, unpublished results.
- [19] A. Fiedler, I. Kretzschmar, D. Schroder, H. Schwarz, Chromium dioxide cation OCrO^+ in the gas phase: structure, electronic states, and the reactivity with hydrogen and hydrocarbons, *J. Am. Chem. Soc.* 118 (1996) 9941-9952.

Tables

Table 1: Chromium polyatomic ion identities. All species except for $^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$ have been isotopically matched to another chromium polyatomic ion to within 5%.

Ion Mass	Ion Identity
81.9363	$^{50}\text{Cr}^{16}\text{O}_2^+$
82.9441	$^{50}\text{Cr}^{16}\text{O}_2\text{H}^+$
83.9303	$^{52}\text{Cr}^{16}\text{O}_2^+$
83.9519	$^{50}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
84.9305	$^{53}\text{Cr}^{16}\text{O}_2^+$
84.9382	$^{52}\text{Cr}^{16}\text{O}_2\text{H}^+$
85.9287	$^{54}\text{Cr}^{16}\text{O}_2^+$
85.9383	$^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$
85.9460	$^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
86.9365	$^{54}\text{Cr}^{16}\text{O}_2\text{H}^+$
86.9461	$^{53}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
87.9444	$^{54}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
97.9312	$^{50}\text{Cr}^{16}\text{O}_3^+$
98.9390	$^{50}\text{Cr}^{16}\text{O}_3\text{H}^+$
99.9252	$^{52}\text{Cr}^{16}\text{O}_3^+$
99.9469	$^{50}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
100.9331	$^{52}\text{Cr}^{16}\text{O}_3\text{H}^+$
101.9332	$^{53}\text{Cr}^{16}\text{O}_3\text{H}^+$
101.9409	$^{52}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
102.9314	$^{54}\text{Cr}^{16}\text{O}_3\text{H}^+$
102.9410	$^{53}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
103.9393	$^{54}\text{Cr}^{16}\text{O}_3\text{H}_2^+$

Table 2: Energies of states of CrOH^+ and CrO^+

	CrOH^+ (ground)	CrOH^+ (excited)	CrO^+
Dissociation Energy to CrO^+ (eV)	1.105	0.410	
Ground State Degeneracy	3	3	8
Vibrational Modes	3	3	1
Vibrational Energies (cm^{-1})	402.42	425.36	585.21
	748.55	745.77	
	3854.6	3856.0	
Rotational Constant x (cm^{-1})	49.1	50.1	0.507
Rotational Constant y (cm^{-1})	0.386	0.385	
Rotational Constant z (cm^{-1})	0.383	0.382	

Table 3: Energies of states of CrO_2^+

	CrO_2^+	CrO_2^+
Dissociation Energy to CrO^+ (eV)	3.711	3.711
Electronic State	A_1	B_2
Ground State Degeneracy	2	2
Vibrational Modes	3	3
Vibrational Energies (cm^{-1})	348.21	357.24
	959.14	975.83
	959.14	1022.3
Rotational Constant x (cm^{-1})	1.09	1.15
Rotational Constant y (cm^{-1})	0.298	0.296
Rotational Constant z (cm^{-1})	0.234	0.236

Figures

Figure 1a: Medium resolution spectrum of 50 ppm Cr(III) in 1% nitric acid. Asterisk denotes ions for which high resolution resolves two separate ions.

Peak Centroid	Integrated Intensity	Assigned Identity
83.9328	2258839	$^{52}\text{Cr}^{16}\text{O}_2^+$
83.9520	9892	$^{50}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
84.9383	744414	* $^{53}\text{Cr}^{16}\text{O}_2^+$
85.9315	83626	$^{54}\text{Cr}^{16}\text{O}_2^+$
85.9474	168948	* $^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
86.9324	35317	$^{54}\text{Cr}^{16}\text{O}_2\text{H}^+$
86.9487	21353	$^{53}\text{Cr}^{16}\text{O}_2\text{H}_2^+$

Figure 1b: High resolution spectrum of 50 ppm Cr(III) in 1% nitric acid.

Peak Centroid	Integrated Intensity	Assigned Identity
83.9392	634716	$^{52}\text{Cr}^{16}\text{O}_2^+$
83.9605	12636	$^{50}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
84.9399	82677	$^{53}\text{Cr}^{16}\text{O}_2^+$
84.9473	172182	$^{52}\text{Cr}^{16}\text{O}_2\text{H}^+$
85.9385	19322	$^{54}\text{Cr}^{16}\text{O}_2^+$
85.9476	30339	$^{53}\text{Cr}^{16}\text{O}_2\text{H}^+$
85.9555	237313	$^{52}\text{Cr}^{16}\text{O}_2\text{H}_2^+$
86.9464	5956	$^{54}\text{Cr}^{16}\text{O}_2\text{H}^+$
86.9557	28468	$^{53}\text{Cr}^{16}\text{O}_2\text{H}_2^+$

Figure 2a: Medium resolution spectrum of 50 ppm Cr(III) in 1% nitric acid. Asterisk denotes ions for which high resolution resolves two separate ions.

Peak Centroid	Integrated Intensity	Assigned Identity
99.9302	49565	$^{52}\text{Cr}^{16}\text{O}_3^+$
99.9510	12796	$^{50}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
100.9380	1889377	$^{52}\text{Cr}^{16}\text{O}_3\text{H}^+$
101.9426	467113	* $^{53}\text{Cr}^{16}\text{O}_3\text{H}^+$
102.9406	96063	* $^{54}\text{Cr}^{16}\text{O}_3\text{H}^+$

Figure 2b: High resolution spectrum of 50 ppm Cr(III) in 1% nitric acid. Y axis has been limited to allow for visualization of less intense peaks.

Peak Centroid	Integrated Intensity	Assigned Identity
99.9334	28931	$^{52}\text{Cr}^{16}\text{O}_3^+$
99.9546	6619	$^{50}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
100.9415	1181888	$^{52}\text{Cr}^{16}\text{O}_3\text{H}^+$
101.9421	136916	$^{53}\text{Cr}^{16}\text{O}_3\text{H}^+$
101.9496	130130	$^{52}\text{Cr}^{16}\text{O}_3\text{H}_2^+$
102.9405	35717	$^{54}\text{Cr}^{16}\text{O}_3\text{H}^+$
102.9492	19241	$^{53}\text{Cr}^{16}\text{O}_3\text{H}_2^+$

Figure 3: CrO_xH_y^+ ion intensities. All ions shown are adjusted for mass bias and intensities have been normalized to CrO^+ .

Figure 4: Comparison of ion abundances from different oxidation states of chromium. All ions have been adjusted for mass bias and concentration.

Figure 5: Comparison of different metals. Adjustments have been made for mass bias, concentration, and isotopic abundance. Cr has a value of 1.0 for each ion shown, and all other elements are normalized to Cr.

CHAPTER 5. GENERAL CONCLUSIONS

Polyatomic ions have been a known source of analyte interference in inductively coupled plasma – mass spectrometry since the early days of the technique. Explicit identification is the first aspect in studying any ion in ICP-MS, but one that is often glossed over, particularly in the discussion of polyatomic ions. In the work presented here, many polyatomic ions have been definitively identified. The use of high resolution, accurate m/z measurements, measured mass defects at each magnet mass setting, isotope ratios, and the addition of specialized solvents allows for proposed ion identities to be confirmed.

Calculating the gas kinetic temperature to evaluate a dissociation constant for a proposed ion pair lends insight into the formation or removal process for the polyatomic ion. Common polyatomic ions such as NO^+ and H_2O^+ have been investigated, as well as less studied ions such as N_2H^+ , COH_x^+ , and CrO_xH_y^+ . Some of these ions have not been previously identified in the literature. Changes in cone material and extraction voltage have been studied to determine their impact on polyatomic ion formation. Additionally, different models of neutral gas density have been employed to represent different positions and conditions in the interface region.

In the future, an instrument interface region could be redesigned to reduce the formation of some of the polyatomic ions studied here. Cone angle, spacing, and material should be explored further, as these studies would only require minor adjustments to existing instruments. Polyatomic ions containing carbon, nitrogen, and hydrogen are prevalent in the mass spectrum, and further knowledge about these ions would improve the determination of aluminum, phosphorous, and sulfur, three elements for which ICP-MS faces some challenges.

The research yet to be done on polyatomic ions in ICP-MS is boundless, as they exist in every region of the mass spectrum and seem to play a role in every analysis. My research has not only brought insight to the polyatomic ion issue, but has prepared me well for a wide variety of ICP-MS work that I hope to execute in the future. I look forward to the challenges this issue presents, and I feel confident that strides will continue to be made in the study of polyatomic ions by high resolution ICP-MS.

ACKNOWLEDGEMENTS

This work was performed at Ames Laboratory under Contract No. W-7405-Eng-82 with the U.S. Department of Energy. The United States government has assigned the DOE Report number IS-T 2357 to this thesis. This work was funded by the National Science Foundation with program assistance from the Institute for Physical Research and Technology.

The work discussed in this text could not have been done without the guidance and support of Professor R. S. Houk. The camaraderie and assistance of past and present Houk group members has been much appreciated. I also thank Dr. Timothy Dudley for his work on calculating the energies of many of the ions discussed. I was honored to be awarded the Fassel fellowship for the 2005-2006 academic year, and for that I thank Mary K. Fassel. Additionally, I wish to show appreciation to Andrew Wade Ferguson, for being my strongest advocate.