GEOMETRICAL INTERPRETATION OF NONLINEARITIES FROM A CYLINDRICAL PICK-UP*

R. Miyamoto, S. E. Kopp, University of Texas at Austin, Austin, TX 78712, USA
A. Jansson, M. J. Syphers, Fermi National Accelerator Laboratory, Batavia, IL 60510, USA

Abstract

In many accelerators, cylindrical pick-ups are used to measure transverse beam positions. Theoretically, signals from these pick-ups are related to infinite power series of the beam position but, in practice, only finite number of terms are considered and the position measurements degrade when a beam is far from the center of a pick-up. This paper shows there is actually a simple geometrical relation between a beam position and induced signals. With help of the geometrical relation, the beam position can be written in a compact function of signals. The paper is concluded with numerical simulations and a test to show this geometry based expression can calculate a beam position better than the conventional methods.

INTRODUCTION

Pick-up detectors are used as transverse beam position monitors (BPMs) in many accelerators. They typically have two or four conducting plates. A beam position is determined from induced image currents on the conducting plates when a beam passes the pick-up. If a beam is far from the center of a pick-up, nonlinearity grows in the relation between the beam position and signals. Then, position measurements can get worse since calibrations get harder. This could be a problem when measurements require high accuracy or the design trajectory is far from the center of a pick-up.

For a cylindrical pick-up, there actually exists a simple geometrical relation between a beam position and induced signals. From the relation, the beam position can be written in a compact form of signals and calculated with much higher accuracy than conventional methods.

GEOMETRY OF A BEAM POSITION AND SIGNALS IN A CYLINDRICAL PICK-UP

Fig 1 shows the cross section of a typical cylindrical pick-up with two conduction plate. For an infinitesimally small beam, the induced current on a conducting plate $\pm$ can be calculated by solving a Poisson problem [1]:

$$\frac{I_{\pm}}{I_b} = \frac{\phi}{2\pi} + \frac{1}{\pi} \sum_{n=1}^{\infty} \frac{2}{n} \left(\frac{\pm r}{b} \right)^n \cos(n\theta) \sin\left(\frac{n\phi}{2}\right), \quad (1)$$

where I_b is the beam current, (r, θ) is the beam position, and b and ϕ are the radius and opening angle of the con-

Figure 1: Cross section of a cylindrical pick-up with two conducting plates. The induced current $I_{\pm}$ on a plate is determined by the beam position (r, θ) and the radius and opening angle of the plates b and ϕ . As explained in the text, there is a simple relation between the current $I_{\pm}$ and angle $\psi_{\pm}$, where $\psi_{\pm}(r, \theta) \equiv \chi(\pm r, \theta) + \chi(\pm r, -\theta)$.

ducting plates. There is a compact form for the infinite sum in Eq 1 and Eq 1 reduces to a following simple form:

$$\frac{I_{\pm}}{I_b} = \frac{\phi}{2\pi} + \frac{\psi_{\pm}}{\pi}, \quad (2)$$

where

$$\psi_{\pm}(r, \theta) \equiv \chi(\pm r, \theta) + \chi(\pm r, -\theta) \quad (3)$$

$$\chi(r, \theta) = \arctan \frac{r \sin(\phi/2 + \theta)}{b - r \cos(\phi/2 + \theta)}. \quad (4)$$

The angle $\chi(\pm r, \pm\theta)$ has a geometrical interpretation seen in Fig 1. From Eq 2, measuring the current $I_{\pm}$ is equivalent to measuring the angle $\psi_{\pm}$.

Consider a circle $C_{\pm}$ which goes through the beam position and the upper and lower edges of the conducting plate $\pm$ (see Fig 2). When a beam is on the circle $C_{\pm}$, the angle $\psi_{\pm}$ remains the same and so does the current $I_{\pm}$. Hence, measuring the current $I_{\pm}$ is also equivalent to determining the circle $C_{\pm}$ and the calculation of an intersection between two circles C_+ and C_- determines the beam position¹:

$$x = \frac{(b/2) \sin \phi \sin(\pi \Delta / I_b)}{\cos(\phi/2) \cos(\pi \Delta / I_b) - \cos(\pi \Sigma / I_b + \phi/2)}, \quad (5)$$

¹The orthogonal displacement can be also determined up to the sign.

* Work supported by Fermilab Research Alliance, LLC under Contract No. DE-AC02-07CH11359 with the United States Department of Energy.

Figure 2: Circles associated to the signals. The circle C_+ goes through the beam position P and the upper and lower edges of the conducting plates $+$, Q and R . From a property of a circle, when a beam is on the circle C_+ , the angle $\angle QPR = \phi + \psi_+$ remains the same and so does the current I_+ . The same argument can be made for the circle C_- , angle $\angle SPT = \phi + \psi_-$, and current I_- . Measuring the signal $I_{\pm}$, measuring the angle $\psi_{\pm}$, and determining the circle $C_{\pm}$ are all equivalent. The beam position is determined by calculating an intersection of C_+ and C_- .

where $\Delta \equiv I_+ - I_-$ and $\Sigma \equiv I_+ + I_-$ are difference and sum signals. Since the beam current I_b cannot be measured with a pick-up, an approximation must be made. For pick-ups with two and four conducting plates, the beam current can be approximated by $I_2 \equiv (\phi/\pi)\Sigma$ and $I_4 \equiv (2\phi/\pi)\Sigma_4$, where Σ_4 is the total induced currents on four conducting plates. By replacing the total current I_b with I_2 or I_4 in Eq 5, the beam position can be calculated from the signals:

$$X_j \equiv \frac{(b/2) \sin \phi \sin(\pi \Delta / I_j)}{\cos(\phi/2) \cos(\pi \Delta / I_j) - \cos(\pi \Sigma / I_j + \phi/2)}, \quad (6)$$

where X_j ($j = 2$ or 4) is the calculated beam position based on this equation when the pick-up has two or four conducting plates. In the following, the upper and lower cases X and x are used as measured and true beam positions.

SIMULATIONS AND A TEST OF NONLINEARITIES IN A TEVATRON BPM

Before thinking about the accuracy of Eq 6, this section discusses the accuracy of the conventional position measurements. As an example, a Tevatron type BPM with $b = 35$ mm and $\phi = 110$ deg is considered [1]. Conventionally, the difference over sum signal Δ/Σ is used to calculate a beam position:

$$\frac{\Delta}{\Sigma} = \frac{4 \sin(\phi/2)}{b\phi} x + (\text{higher orders of } x \text{ and } y). \quad (7)$$

Notice the beam current I_b is not in this equation either.

In the lowest order, the beam position is determined by

Figure 3: The ratio between the calculated and true beam positions when only the lowest order of the difference over sum signal is considered. Simulations are done for a Tevatron type BPM: two conducting plates, $b = 35$ mm and $\phi = 110$ deg. Measurements get worse when the beam is off center in either of transverse directions.

simply ignoring the all higher order terms in Eq 7:

$$X_0 \equiv \frac{b\phi}{4 \sin(\phi/2)} \frac{\Delta}{\Sigma}. \quad (8)$$

Fig 3 shows how the position measurement based on this equation degrades if the beam is off center in either of transverse directions².

When the orthogonal displacement is zero ($y = 0$), the Eq 7 can be numerically solved for x and the beam position is accurately given³. Even when the orthogonal displacement is not zero, the solution of Eq 7 with the assumption of $y = 0$ gives a better position measurement than Eq 8.

²Similar simulations are also seen in [2].

³It corresponds to solve the true beam position x from the measured beam position X_0 based on the curve with $y = 0$ in Fig 3.

Figure 4: The ratio between the calculated and true beam positions X_1 and x . Here, X_1 is the numerical solution of Eq 7 when the zero orthogonal displacement is assumed. Conditions of the simulations are the same as Fig 3. Compared to Fig 3, the measurements are better for small orthogonal displacements but no improvement is made for large orthogonal displacements.

Figure 5: Comparison of measured horizontal positions when vertical positions are different. A beam is moved in both planes by using local four bumps (~ 1 mm/bump) at one horizontal BPM. Measurements are roughly following the model predictions but distributions are large.

Let X_1 be the beam position calculated in this way⁴. Fig 4 shows the ratio between the measured position X_1 and the true position. Since X_1 is perfectly accurate for zero orthogonal displacement, compared to Fig 3, position measurements are better for small orthogonal displacements. On the other hand, accuracies remain the same (or even get worse) for large orthogonal displacements.

Measurements using the present Tevatron BPM system are also based on the solution of Eq 7 assuming zero orthogonal displacement, X_1 . By using local four bumps, a beam is moved in both transverse planes at one BPM and changes in the position measurements are observed⁵. Fig 5 shows differences of the measured horizontal positions when vertical positions are different. As expected from Fig 4, the measured horizontal position gets larger when the vertical displacement increases.

⁴Mathematically, X_1 is the solution of $\frac{\Delta}{\Sigma}(x, y) = \frac{\Delta}{\Sigma}(X_1, 0)$.

⁵An existing algorithm is used for data acquisitions [3].

Figure 6: Simulation of the beam position determined from the geometry based equation X_2 for pick-ups with two conducting plates. The orthogonal displacement is fixed to $b/2$ and each line represents a different opening angle ϕ . When $80 < \phi < 90$ deg, the accuracies are better than 3%.

Figure 7: Simulation of the beam position determined from the geometry based equation X_4 for pick-ups with four conducting plates. The orthogonal displacement is fixed to $b/2$ and each line represents a different opening angle ϕ . The accuracies are better than 1% when $\phi > 75$ deg.

SIMULATIONS OF THE GEOMETRY BASED EQUATIONS X_2 AND X_4

The geometry based equation Eq 6 can calculate a beam position better than conventional methods. Figs 6 and 7 show errors of the calculated positions from X_2 and X_4 when the orthogonal displacement is $b/2$. Here, different lines represent different opening angles. In Fig 4, the line of $y = 16$ mm $\lesssim b/2$ shows more than 10% error. Whereas, the errors of X_2 and X_4 are on the order of a few percent or better for a certain range of the opening angle.

Since the beam current is approximated better for pick-ups with four conducting plates, X_4 has better accuracy than X_2 in general. Figs 6 and 7 indicate the optimum opening angle is about 85 deg for both X_2 and X_4 . It is because $\partial X_j / \partial I_j$ has the minimum around 85 deg and the effect of the beam current approximation is minimized.

CONCLUSION

For a cylindrical pick-up, there is a geometrical relation between a beam position and induced signals and it allows beam position to be written in a compact form of signals. Compared to the conventional methods, the geometry based equation can calculate a beam position better, especially when the orthogonal displacement is large. The optimum opening angle for this equation is about 85 deg.

REFERENCES

- [1] R. E. Shafer, "Characteristics of Directional Coupler Beam Position Monitors", *IEEE. Trans. Nucl. Sci.* NS-32, p.1933 (1985).
- [2] R. K. Kutschke, "Third Order Calculation of Beam Position", Fermilab Report Beams-doc-1893-v3, 2005.
- [3] M. A. Martens, "Position Scan of Tev VPA33 BPM with Upgraded Electronic", Fermilab Report Beams-doc-1516-v1, 2004.